Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREY,ROBERT,P
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: MELIUS,DALTON
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 12/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: BUCK

DATE/TIME OF TEST: 01/02/2018 18:47:53

Diagnostics Check		OK	18:41
Air Blank		0.000	18:42
Control Test		0.080	18:42
Air Blank		0.000	18:43
Subject Sample	# 1	0.000	18:43
Breath Volume		2.144	18:43
Air Blank		0.000	18:44
Air Blank		0.000	18:46
Subject Sample	# 2	0.000	18:46
Breath Volume		1.976	18:46
Air Blank		0.000	18:46
Control Test		0.081	18:47
Air Blank		0.000	18:47
Diagnostics Check		OK	18:47

Data Download Date/Time: 01/30/2018 14:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREY,ROBERT,P
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: HOLLCROFT,D
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 12/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: BREAUX

DATE/TIME OF TEST: 01/10/2018 21:32:49

Diagnostics Check		OK	21:26
Air Blank		0.000	21:26
Control Test		0.081	21:27
Air Blank		0.000	21:27
Subject Sample	# 1	0.295	21:28
Breath Volume		2.425	21:28
Air Blank		0.000	21:28
Air Blank		0.000	21:30
Subject Sample	# 2	0.286	21:31
Breath Volume		3.804	21:31
Air Blank		0.000	21:31
Control Test		0.079	21:32
Air Blank		0.000	21:32
Diagnostics Check		OK	21:32

Data Download Date/Time: 01/30/2018 14:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREY,ROBERT,P
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: HOLLCROFT,D
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 12/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: BIDEZ

DATE/TIME OF TEST: 01/16/2018 23:31:14

Diagnostics Check		OK	23:24
Air Blank		0.000	23:24
Control Test		0.081	23:25
Air Blank		0.000	23:25
Subject Sample	# 1	0.000	23:26
Breath Volume		2.449	23:26
Air Blank		0.000	23:26
Air Blank		0.000	23:28
Subject Sample	# 2	0.000	23:29
Breath Volume		1.472	23:29
Air Blank		0.000	23:30
Control Test		0.081	23:30
Air Blank		0.000	23:31
Diagnostics Check		OK	23:31

Data Download Date/Time: 01/30/2018 14:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORGAN,D,
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: MORGAN,D
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 12/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 14:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:21

SUBJECT'S LAST NAME: HOUSSANY

DATE/TIME OF TEST: 01/18/2018 15:51:49

Diagnostics Check		OK	15:45
Air Blank		0.000	15:46
Control Test		0.081	15:46
Air Blank		0.000	15:47
Subject Sample	# 1	0.000	15:47
Breath Volume		1.601	15:47
Air Blank		0.000	15:48
Air Blank		0.000	15:50
Subject Sample	# 2	0.000	15:50
Breath Volume		1.460	15:50
Air Blank		0.000	15:50
Control Test		0.081	15:51
Air Blank		0.000	15:51
Diagnostics Check		OK	15:51

Data Download Date/Time: 01/30/2018 14:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,GREG,C
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: ABROE,NICHOLAS
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 12/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 12:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:12

SUBJECT'S LAST NAME: POWELL

DATE/TIME OF TEST: 01/22/2018 12:57:52

Diagnostics Check		OK	12:51
Air Blank		0.000	12:52
Control Test		0.081	12:52
Air Blank		0.000	12:53
Subject Sample	# 1	0.000	12:53
Breath Volume		2.238	12:53
Air Blank		0.000	12:54
Air Blank		0.000	12:56
Subject Sample	# 2	0.000	12:56
Breath Volume		2.222	12:56
Air Blank		0.000	12:56
Control Test		0.080	12:57
Air Blank		0.000	12:57
Diagnostics Check		OK	12:57

Data Download Date/Time: 01/30/2018 14:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHILKE,CARL,E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SCHULTZ,KEVIN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:01

SUBJECT'S LAST NAME: TORMEY

DATE/TIME OF TEST: 01/01/2018 20:02:31

Diagnostics Check		OK	19:55
Air Blank		0.000	19:56
Control Test		0.081	19:56
Air Blank		0.000	19:57
Subject Sample	# 1	0.224	19:57
Breath Volume		2.269	19:57
Air Blank		0.000	19:58
Air Blank		0.000	20:00
Subject Sample	# 2	0.227	20:00
Breath Volume		1.296	20:00
Air Blank		0.000	20:01
Control Test		0.081	20:01
Air Blank		0.000	20:02
Diagnostics Check		OK	20:02

Data Download Date/Time: 01/10/2018 10:43

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHILKE,CARL,E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SCHILKE,CARL
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: PRATT

DATE/TIME OF TEST: 01/01/2018 20:11:44

Diagnostics Check		OK	20:05
Air Blank		0.000	20:05
Control Test		0.081	20:06
Air Blank		0.000	20:06
Subject Sample	# 1	0.124	20:07
Breath Volume		2.417	20:07
Air Blank		0.000	20:07
Air Blank		0.000	20:09
Subject Sample	# 2	0.120	20:10
Breath Volume		2.636	20:10
Air Blank		0.000	20:10
Control Test		0.081	20:11
Air Blank		0.000	20:11
Diagnostics Check		OK	20:11

Data Download Date/Time: 01/10/2018 10:43

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHILKE,CARL,E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: YOUNGER,M
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 01/01/2018 21:09:12

Diagnostics Check		OK	21:02
Air Blank		0.000	21:03
Control Test		0.081	21:03
Air Blank		0.000	21:03
Subject Sample	# 1	0.281	21:04
Breath Volume		3.160	21:04
Air Blank		0.000	21:05
Air Blank		0.000	21:06
Subject Sample	# 2	0.286	21:07
Breath Volume		3.144	21:07
Air Blank		0.000	21:08
Control Test		0.080	21:08
Air Blank		0.000	21:09
Diagnostics Check		OK	21:09

Data Download Date/Time: 01/10/2018 10:43

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WITT,KELLY,
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: WITT,KELLY
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: COBURN

DATE/TIME OF TEST: 01/02/2018 00:28:28

Diagnostics Check		OK	00:22
Air Blank		0.000	00:22
Control Test		0.082	00:23
Air Blank		0.000	00:23
Subject Sample	# 1	0.000	00:24
Breath Volume		3.062	00:24
Air Blank		0.000	00:24
Air Blank		0.000	00:26
Subject Sample	# 2	0.000	00:27
Breath Volume		2.398	00:27
Air Blank		0.000	00:27
Control Test		0.081	00:27
Air Blank		0.000	00:28
Diagnostics Check		OK	00:28

Data Download Date/Time: 01/10/2018 10:43

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAYLOR,MAURICE,E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: NAYLOR,MAURICE
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: BLOSE

DATE/TIME OF TEST: 01/03/2018 02:27:48

Diagnostics Check		OK	02:18	
Air Blank		0.000	02:19	
Control Test		0.081	02:19	
Air Blank		0.000	02:19	
Subject Sample	# 1	0.177	02:21	
Breath Volume		1.253	02:21	
Air Blank		0.000	02:22	
Air Blank		0.000	02:23	
Subject Sample	# 2		02:26	Subject Test Refused
Breath Volume		0.000	02:26	
Air Blank		0.000	02:26	
Control Test		0.080	02:27	
Air Blank		0.000	02:27	
Diagnostics Check		OK	02:27	

Data Download Date/Time: 01/10/2018 10:43

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JUAN,AVENDANO,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: MCBRIDE,TIM
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: ISNORD

DATE/TIME OF TEST: 01/05/2018 04:05:40

Diagnostics Check		OK	03:58
Air Blank		0.000	03:58
Control Test		0.082	03:58
Air Blank		0.000	03:59
Subject Sample	# 1	0.165	04:00
Breath Volume		1.382	04:00
Air Blank		0.000	04:01
Air Blank		0.000	04:03
Subject Sample	# 2	0.165	04:04
Breath Volume		1.546	04:04
Air Blank		0.000	04:04
Control Test		0.080	04:05
Air Blank		0.000	04:05
Diagnostics Check		OK	04:05

Data Download Date/Time: 01/10/2018 10:43

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,YOEL,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: RONGISH,SHERRY
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 11:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:41

SUBJECT'S LAST NAME: GAFFORD

DATE/TIME OF TEST: 01/06/2018 13:13:00

Diagnostics Check		OK	13:05
Air Blank		0.000	13:06
Control Test		0.082	13:06
Air Blank		0.000	13:07
Subject Sample	# 1	0.128	13:07
Breath Volume		3.859	13:07
Air Blank		0.000	13:08
Air Blank		0.000	13:10
Subject Sample	# 2	0.131	13:11
Breath Volume		3.812	13:11
Air Blank		0.000	13:12
Control Test		0.082	13:12
Air Blank		0.000	13:12
Diagnostics Check		OK	13:13

Data Download Date/Time: 01/10/2018 10:43

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WITT,KELLY,
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: LOUIS,RENET
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 13:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:35

SUBJECT'S LAST NAME: NOLASCO

DATE/TIME OF TEST: 01/06/2018 17:14:10

Diagnostics Check		OK	17:08
Air Blank		0.000	17:08
Control Test		0.081	17:08
Air Blank		0.000	17:09
Subject Sample	# 1	0.192	17:09
Breath Volume		3.031	17:09
Air Blank		0.000	17:10
Air Blank		0.000	17:12
Subject Sample	# 2	0.191	17:12
Breath Volume		1.492	17:12
Air Blank		0.000	17:13
Control Test		0.081	17:13
Air Blank		0.000	17:14
Diagnostics Check		OK	17:14

Data Download Date/Time: 01/10/2018 10:43

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAYLOR,MAURICE,E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LINDHOLM,JOHN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: COSME GARCIA

DATE/TIME OF TEST: 01/07/2018 03:06:34

Diagnostics Check		OK	02:55
Air Blank		0.000	02:55
Control Test		0.081	02:55
Air Blank		0.000	02:56
Subject Sample	# 1	0.188	02:59
Breath Volume		1.164	02:59
Air Blank		0.000	02:59
Air Blank		0.000	03:01
Subject Sample	# 2	0.185	03:04
Breath Volume		1.652	03:04
Air Blank		0.000	03:05
Control Test		0.078	03:05
Air Blank		0.000	03:06
Diagnostics Check		OK	03:06

Data Download Date/Time: 01/10/2018 10:43

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHILKE,CARL,E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: HIDEG,JARED
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: CAO

DATE/TIME OF TEST: 01/07/2018 23:38:19

Diagnostics Check		OK	23:30
Air Blank		0.000	23:30
Control Test		0.081	23:31
Air Blank		0.000	23:31
Subject Sample	# 1	0.173	23:32
Breath Volume		2.738	23:32
Air Blank		0.000	23:33
Air Blank		0.000	23:35
Subject Sample	# 2	0.170	23:36
Breath Volume		2.535	23:36
Air Blank		0.000	23:37
Control Test		0.080	23:37
Air Blank		0.000	23:38
Diagnostics Check		OK	23:38

Data Download Date/Time: 01/10/2018 10:43

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHILKE,CARL,E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: STREET,JASON
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: WEYH

DATE/TIME OF TEST: 01/08/2018 00:24:25

Diagnostics Check		OK	00:18
Air Blank		0.000	00:18
Control Test		0.080	00:18
Air Blank		0.000	00:19
Subject Sample	# 1	0.106	00:19
Breath Volume		1.105	00:19
Air Blank		0.000	00:20
Air Blank		0.000	00:22
Subject Sample	# 2	0.102	00:22
Breath Volume		1.386	00:22
Air Blank		0.000	00:23
Control Test		0.080	00:23
Air Blank		0.000	00:24
Diagnostics Check		OK	00:24

Data Download Date/Time: 01/10/2018 10:43

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANPELT,J,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: DEPRIEST,R
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: HIGHMAN

DATE/TIME OF TEST: 01/08/2018 03:36:27

Diagnostics Check		OK	03:29
Air Blank		0.000	03:29
Control Test		0.081	03:30
Air Blank		0.000	03:30
Subject Sample	# 1	0.171	03:31
Breath Volume		2.464	03:31
Air Blank		0.000	03:32
Air Blank		0.000	03:33
Subject Sample	# 2	0.178	03:34
Breath Volume		2.644	03:34
Air Blank		0.000	03:35
Control Test		0.081	03:35
Air Blank		0.000	03:36
Diagnostics Check		OK	03:36

Data Download Date/Time: 01/10/2018 10:43

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANPELT,J,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: HIDEG,J
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: REYES ESPINOSA

DATE/TIME OF TEST: 01/12/2018 04:27:20

Diagnostics Check		OK	04:20
Air Blank		0.000	04:21
Control Test		0.081	04:21
Air Blank		0.000	04:22
Subject Sample	# 1	0.146	04:22
Breath Volume		3.140	04:22
Air Blank		0.000	04:23
Air Blank		0.000	04:25
Subject Sample	# 2	0.151	04:25
Breath Volume		3.015	04:25
Air Blank		0.000	04:26
Control Test		0.081	04:26
Air Blank		0.000	04:27
Diagnostics Check		OK	04:27

Data Download Date/Time: 02/14/2018 08:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNCAN,N,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: DUNCAN,N
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: BAUTISTA SANCHEZ

DATE/TIME OF TEST: 01/13/2018 02:02:26

Diagnostics Check		OK	01:55
Air Blank		0.000	01:55
Control Test		0.081	01:55
Air Blank		0.000	01:56
Subject Sample	# 1	0.225	01:57
Breath Volume		2.156	01:57
Air Blank		0.000	01:58
Air Blank		0.000	02:00
Subject Sample	# 2	0.217	02:00
Breath Volume		2.671	02:00
Air Blank		0.000	02:01
Control Test		0.079	02:01
Air Blank		0.000	02:02
Diagnostics Check		OK	02:02

Data Download Date/Time: 02/14/2018 08:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROMAN,ANTHONY,C
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: ELSAID,M
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 13:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: HILLEARY

DATE/TIME OF TEST: 01/13/2018 14:42:38

Diagnostics Check		OK	14:36
Air Blank		0.000	14:36
Control Test		0.082	14:36
Air Blank		0.000	14:37
Subject Sample	# 1	0.238	14:37
Breath Volume		4.363	14:37
Air Blank		0.000	14:38
Air Blank		0.000	14:40
Subject Sample	# 2	0.246	14:40
Breath Volume		4.460	14:40
Air Blank		0.000	14:41
Control Test		0.080	14:42
Air Blank		0.000	14:42
Diagnostics Check		OK	14:42

Data Download Date/Time: 02/14/2018 08:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNCAN,N,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: DUNCAN,N
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:34

SUBJECT'S LAST NAME: MOEHRING

DATE/TIME OF TEST: 01/13/2018 22:02:54

Diagnostics Check		OK	21:56
Air Blank		0.000	21:57
Control Test		0.082	21:57
Air Blank		0.000	21:57
Subject Sample	# 1	0.083	21:58
Breath Volume		2.371	21:58
Air Blank		0.000	21:59
Air Blank		0.000	22:00
Subject Sample	# 2	0.085	22:01
Breath Volume		2.285	22:01
Air Blank		0.000	22:01
Control Test		0.081	22:02
Air Blank		0.000	22:02
Diagnostics Check		OK	22:02

Data Download Date/Time: 02/14/2018 08:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVENDANO,JUAN,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: BONSALL,MATTHEW
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: CLAPPER

DATE/TIME OF TEST: 01/15/2018 01:08:27

Diagnostics Check		OK	01:01
Air Blank		0.000	01:01
Control Test		0.080	01:02
Air Blank		0.000	01:02
Subject Sample	# 1	0.218	01:03
Breath Volume		2.457	01:03
Air Blank		0.000	01:04
Air Blank		0.000	01:05
Subject Sample	# 2	0.219	01:06
Breath Volume		2.121	01:06
Air Blank		0.000	01:07
Control Test		0.081	01:07
Air Blank		0.000	01:08
Diagnostics Check		OK	01:08

Data Download Date/Time: 02/14/2018 08:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHILKE,CARL,E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: CAIL,MATTHEW
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: VAN ZANDT

DATE/TIME OF TEST: 01/17/2018 03:49:49

Diagnostics Check		OK	03:43
Air Blank		0.000	03:43
Control Test		0.082	03:44
Air Blank		0.000	03:44
Subject Sample	# 1	0.218	03:45
Breath Volume		1.578	03:45
Air Blank		0.000	03:45
Air Blank		0.000	03:47
Subject Sample	# 2	0.218	03:48
Breath Volume		1.636	03:48
Air Blank		0.000	03:48
Control Test		0.081	03:49
Air Blank		0.000	03:49
Diagnostics Check		OK	03:49

Data Download Date/Time: 02/14/2018 08:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMYRNIOS,MANUEL,J
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: SMYRNIOS,MANUEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 15:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: DENMAN

DATE/TIME OF TEST: 01/19/2018 17:42:51

Diagnostics Check		OK	17:36
Air Blank		0.000	17:36
Control Test		0.081	17:37
Air Blank		0.000	17:37
Subject Sample	# 1	0.253	17:38
Breath Volume		3.089	17:38
Air Blank		0.000	17:39
Air Blank		0.000	17:40
Subject Sample	# 2	0.250	17:41
Breath Volume		2.886	17:41
Air Blank		0.000	17:41
Control Test		0.079	17:42
Air Blank		0.000	17:42
Diagnostics Check		OK	17:42

Data Download Date/Time: 02/14/2018 08:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOULD,LEONARD,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GOULD,LEONARD
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: STOVER

DATE/TIME OF TEST: 01/19/2018 23:16:58

Diagnostics Check		OK	23:10
Air Blank		0.000	23:10
Control Test		0.082	23:11
Air Blank		0.000	23:11
Subject Sample	# 1	0.176	23:12
Breath Volume		3.054	23:12
Air Blank		0.000	23:13
Air Blank		0.000	23:14
Subject Sample	# 2	0.169	23:15
Breath Volume		2.941	23:15
Air Blank		0.000	23:16
Control Test		0.081	23:16
Air Blank		0.000	23:16
Diagnostics Check		OK	23:16

Data Download Date/Time: 02/14/2018 08:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHILKE,CARL,E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SCHILKE,CARL
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:53

SUBJECT'S LAST NAME: GOETZ

DATE/TIME OF TEST: 01/21/2018 18:22:52

Diagnostics Check		OK	18:15
Air Blank		0.000	18:15
Control Test		0.081	18:16
Air Blank		0.000	18:16
Subject Sample	# 1	0.275	18:17
Breath Volume		1.964	18:17
Air Blank		0.000	18:17
Air Blank		0.000	18:19
Subject Sample	# 2	0.266	18:21
Breath Volume		1.843	18:21
Air Blank		0.000	18:21
Control Test		0.080	18:22
Air Blank		0.000	18:22
Diagnostics Check		OK	18:22

Data Download Date/Time: 02/14/2018 08:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,J,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: RIVERA,J
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:19

SUBJECT'S LAST NAME: GILCHRIST

DATE/TIME OF TEST: 01/21/2018 21:49:15

Diagnostics Check		OK	21:42
Air Blank		0.000	21:42
Control Test		0.081	21:43
Air Blank		0.000	21:43
Subject Sample	# 1	0.146	21:44
Breath Volume		1.281	21:44
Air Blank		0.000	21:45
Air Blank		0.000	21:46
Subject Sample	# 2	0.159	21:47
Breath Volume		2.195	21:47
Air Blank		0.000	21:48
Control Test		0.080	21:48
Air Blank		0.000	21:49
Diagnostics Check		OK	21:49

Data Download Date/Time: 02/14/2018 08:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNCAN,N,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: DUNCAN,N
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: FERRELL

DATE/TIME OF TEST: 01/22/2018 23:41:43

Diagnostics Check		OK	23:35
Air Blank		0.000	23:36
Control Test		0.081	23:36
Air Blank		0.000	23:37
Subject Sample	# 1	0.000	23:37
Breath Volume		1.660	23:37
Air Blank		0.000	23:37
Air Blank		0.000	23:39
Subject Sample	# 2	0.000	23:40
Breath Volume		1.265	23:40
Air Blank		0.000	23:40
Control Test		0.080	23:41
Air Blank		0.000	23:41
Diagnostics Check		OK	23:41

Data Download Date/Time: 02/14/2018 08:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHILKE,CARL,E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LINDHOLM,JOHN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: WYSKOCHIL

DATE/TIME OF TEST: 01/24/2018 22:42:34

Diagnostics Check		OK	22:35
Air Blank		0.000	22:36
Control Test		0.082	22:36
Air Blank		0.000	22:37
Subject Sample	# 1	0.192	22:37
Breath Volume		3.476	22:37
Air Blank		0.000	22:38
Air Blank		0.000	22:40
Subject Sample	# 2	0.197	22:40
Breath Volume		2.816	22:40
Air Blank		0.000	22:41
Control Test		0.081	22:41
Air Blank		0.000	22:42
Diagnostics Check		OK	22:42

Data Download Date/Time: 02/14/2018 08:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEYERS,F,E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: NATAL,J
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: MOHR

DATE/TIME OF TEST: 01/25/2018 02:11:56

Diagnostics Check		OK	02:05
Air Blank		0.000	02:05
Control Test		0.081	02:06
Air Blank		0.000	02:06
Subject Sample	# 1	0.245	02:07
Breath Volume		3.210	02:07
Air Blank		0.000	02:08
Air Blank		0.000	02:09
Subject Sample	# 2	0.241	02:10
Breath Volume		2.222	02:10
Air Blank		0.000	02:11
Control Test		0.081	02:11
Air Blank		0.000	02:11
Diagnostics Check		OK	02:11

Data Download Date/Time: 02/14/2018 08:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOULD,LEONARD,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: DAVIS,DANIEL
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 01/26/2018 01:27:28

Diagnostics Check		OK	01:20
Air Blank		0.000	01:21
Control Test		0.081	01:21
Air Blank		0.000	01:21
Subject Sample	# 1	0.140	01:22
Breath Volume		2.664	01:22
Air Blank		0.000	01:23
Air Blank		0.000	01:25
Subject Sample	# 2	0.138	01:25
Breath Volume		2.343	01:25
Air Blank		0.000	01:26
Control Test		0.080	01:26
Air Blank		0.000	01:27
Diagnostics Check		OK	01:27

Data Download Date/Time: 02/14/2018 08:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOULD,LEONARD,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: CALHOUN,KURTIS
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: PLOURDE

DATE/TIME OF TEST: 01/26/2018 03:02:40

Diagnostics Check		OK	02:55
Air Blank		0.000	02:56
Control Test		0.081	02:56
Air Blank		0.000	02:57
Subject Sample	# 1	0.158	02:58
Breath Volume		1.800	02:58
Air Blank		0.000	02:58
Air Blank		0.000	03:00
Subject Sample	# 2	0.165	03:01
Breath Volume		1.230	03:01
Air Blank		0.000	03:01
Control Test		0.081	03:02
Air Blank		0.000	03:02
Diagnostics Check		OK	03:02

Data Download Date/Time: 02/14/2018 08:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVENDANO,JUAN,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: STEVENS,LAUREN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:51

SUBJECT'S LAST NAME: SCHROLL

DATE/TIME OF TEST: 01/26/2018 21:45:12

Diagnostics Check		OK	21:34
Air Blank		0.000	21:34
Control Test		0.081	21:35
Air Blank		0.000	21:35
Subject Sample	# 1	0.242	21:36
Breath Volume		1.140	21:36
Air Blank		0.000	21:36
Air Blank		0.000	21:38
Subject Sample	# 2	0.269	21:39
Breath Volume		1.710	21:39
Air Blank		0.000	21:40
Air Blank		0.000	21:42
Subject Sample	# 3	0.264	21:43
Breath Volume		1.285	21:43
Air Blank		0.000	21:44
Control Test		0.079	21:44
Air Blank		0.000	21:45
Diagnostics Check		OK	21:45

Data Download Date/Time: 02/14/2018 08:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SLAFER,GAGE,R
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: SLAFER,GAGE
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 13:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:28

SUBJECT'S LAST NAME: LAMORTE

DATE/TIME OF TEST: 01/27/2018 15:22:02

Diagnostics Check		OK	15:10	
Air Blank		0.000	15:11	
Control Test		0.081	15:11	
Air Blank		0.000	15:11	
Subject Sample	# 1	0.137	15:15	Volume Not Met
Breath Volume		0.917	15:15	
Air Blank		0.000	15:15	
Air Blank		0.000	15:17	
Subject Sample	# 2		15:20	No Sample Provided
Breath Volume		0.000	15:20	
Air Blank		0.000	15:21	
Control Test		0.077	15:21	
Air Blank		0.000	15:21	
Diagnostics Check		OK	15:22	

Data Download Date/Time: 02/14/2018 08:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOULD,LEONARD,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: FOGLE,FLOYD
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 01/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:02

SUBJECT'S LAST NAME: KIEFFER

DATE/TIME OF TEST: 01/29/2018 19:30:07

Diagnostics Check		OK	19:22
Air Blank		0.000	19:23
Control Test		0.081	19:23
Air Blank		0.000	19:24
Subject Sample	# 1	0.150	19:25
Breath Volume		1.679	19:25
Air Blank		0.000	19:25
Air Blank		0.000	19:27
Subject Sample	# 2	0.156	19:28
Breath Volume		1.609	19:28
Air Blank		0.000	19:29
Control Test		0.080	19:29
Air Blank		0.000	19:30
Diagnostics Check		OK	19:30

Data Download Date/Time: 02/14/2018 08:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,MICHAEL,K
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: REED,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: MANZELLI

DATE/TIME OF TEST: 01/29/2018 22:13:04

Diagnostics Check		OK	22:06
Air Blank		0.000	22:06
Control Test		0.081	22:07
Air Blank		0.000	22:07
Subject Sample	# 1	0.000	22:08
Breath Volume		2.902	22:08
Air Blank		0.000	22:09
Air Blank		0.000	22:11
Subject Sample	# 2	0.000	22:11
Breath Volume		3.371	22:11
Air Blank		0.000	22:12
Control Test		0.081	22:12
Air Blank		0.000	22:12
Diagnostics Check		OK	22:13

Data Download Date/Time: 02/14/2018 08:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JASIUKIEWICZ,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: WHALEN

DATE/TIME OF TEST: 01/01/2018 02:13:48

Diagnostics Check		OK	02:10	
Air Blank		0.000	02:11	
Control Test		0.079	02:11	
Air Blank		0.000	02:12	
Subject Sample	# 1		02:12	Subject Test Refused
Breath Volume		0.000	02:12	
Air Blank		0.000	02:12	
Control Test		0.079	02:13	
Air Blank		0.000	02:13	
Diagnostics Check		OK	02:13	

Data Download Date/Time: 01/05/2018 10:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEA,JOHN,D
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SHEA,JOHN
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 16:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:37

SUBJECT'S LAST NAME: PECKHAM

DATE/TIME OF TEST: 01/03/2018 17:03:38

Diagnostics Check		OK	16:57
Air Blank		0.000	16:58
Control Test		0.080	16:58
Air Blank		0.000	16:58
Subject Sample	# 1	0.000	16:59
Breath Volume		1.984	16:59
Air Blank		0.000	16:59
Air Blank		0.000	17:01
Subject Sample	# 2	0.000	17:02
Breath Volume		2.199	17:02
Air Blank		0.000	17:02
Control Test		0.080	17:03
Air Blank		0.000	17:03
Diagnostics Check		OK	17:03

Data Download Date/Time: 01/05/2018 10:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JASIUKIEWICZ,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: MAY

DATE/TIME OF TEST: 01/05/2018 02:06:25

Diagnostics Check		OK	02:03	
Air Blank		0.000	02:04	
Control Test		0.080	02:04	
Air Blank		0.000	02:04	
Subject Sample	# 1		02:05	Subject Test Refused
Breath Volume		0.000	02:05	
Air Blank		0.000	02:05	
Control Test		0.079	02:05	
Air Blank		0.000	02:06	
Diagnostics Check		OK	02:06	

Data Download Date/Time: 01/05/2018 10:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JASIUKIEWICZ,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 01/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: WHITMORE

DATE/TIME OF TEST: 01/05/2018 22:48:39

Diagnostics Check		OK	22:45	
Air Blank		0.000	22:46	
Control Test		0.080	22:46	
Air Blank		0.000	22:47	
Subject Sample	# 1		22:47	Subject Test Refused
Breath Volume		0.000	22:47	
Air Blank		0.000	22:47	
Control Test		0.081	22:48	
Air Blank		0.000	22:48	
Diagnostics Check		OK	22:48	

Data Download Date/Time: 02/01/2018 14:37

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JASIUKIEWICZ,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 01/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 01/06/2018 23:57:42

Diagnostics Check		OK	23:54	
Air Blank		0.000	23:55	
Control Test		0.080	23:55	
Air Blank		0.000	23:56	
Subject Sample	# 1		23:56	Subject Test Refused
Breath Volume		0.000	23:56	
Air Blank		0.000	23:56	
Control Test		0.080	23:57	
Air Blank		0.000	23:57	
Diagnostics Check		OK	23:57	

Data Download Date/Time: 02/01/2018 14:37

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAZZELLA,JOSEPH,T
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: HEVEL,MATTHEW
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 01/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 14:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:46

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 01/07/2018 15:26:03

Diagnostics Check		OK	15:14	
Air Blank		0.000	15:15	
Control Test		0.079	15:15	
Air Blank		0.000	15:15	
Subject Sample	# 1		15:19	No Sample Provided
Breath Volume		0.000	15:19	
Air Blank		0.000	15:19	
Air Blank		0.000	15:21	
Subject Sample	# 2		15:24	No Sample Provided
Breath Volume		0.000	15:24	
Air Blank		0.000	15:25	
Control Test		0.078	15:25	
Air Blank		0.000	15:25	
Diagnostics Check		OK	15:26	

Data Download Date/Time: 02/01/2018 14:37

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 01/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: HILLHOUSE

DATE/TIME OF TEST: 01/09/2018 00:10:11

Diagnostics Check		OK	00:05	
Air Blank		0.000	00:05	
Control Test		0.079	00:06	
Air Blank		0.000	00:06	
Subject Sample	# 1		00:08	Subject Test Refused
Breath Volume		0.000	00:08	
Air Blank		0.000	00:09	
Control Test		0.081	00:09	
Air Blank		0.000	00:10	
Diagnostics Check		OK	00:10	

Data Download Date/Time: 02/01/2018 14:37

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COUCH,WILLIAM,K
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: COUCH,WILLIAM
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 01/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: DESPER

DATE/TIME OF TEST: 01/09/2018 01:11:46

Diagnostics Check		OK	01:07	
Air Blank		0.000	01:08	
Control Test		0.080	01:08	
Air Blank		0.000	01:08	
Subject Sample	# 1		01:10	Subject Test Refused
Breath Volume		0.000	01:10	
Air Blank		0.000	01:10	
Control Test		0.078	01:11	
Air Blank		0.000	01:11	
Diagnostics Check		OK	01:11	

Data Download Date/Time: 02/01/2018 14:37

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COUCH,WILLIAM,K
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: BRUNER,SEAN
ARRESTING OFFICER AGENCY: Kenneth City PD
LAST AGENCY INSPECTION: 01/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: SHRADER

DATE/TIME OF TEST: 01/10/2018 02:30:42

Diagnostics Check		OK	02:24
Air Blank		0.000	02:24
Control Test		0.079	02:25
Air Blank		0.000	02:25
Subject Sample	# 1	0.000	02:26
Breath Volume		3.097	02:26
Air Blank		0.000	02:26
Air Blank		0.000	02:28
Subject Sample	# 2	0.000	02:29
Breath Volume		1.886	02:29
Air Blank		0.000	02:29
Control Test		0.080	02:30
Air Blank		0.000	02:30
Diagnostics Check		OK	02:30

Data Download Date/Time: 02/01/2018 14:37

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JASIUKIEWICZ,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 01/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 01/12/2018 00:46:47

Diagnostics Check		OK	00:39
Air Blank		0.000	00:39
Control Test		0.079	00:40
Air Blank		0.000	00:40
Subject Sample	# 1	0.103	00:42
Breath Volume		2.628	00:42
Air Blank		0.000	00:43
Air Blank		0.000	00:44
Subject Sample	# 2	0.100	00:45
Breath Volume		1.796	00:45
Air Blank		0.000	00:45
Control Test		0.079	00:46
Air Blank		0.000	00:46
Diagnostics Check		OK	00:46

Data Download Date/Time: 02/01/2018 14:37

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JASIUKIEWICZ,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 01/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 01/14/2018 03:03:19

Diagnostics Check		OK	02:57	
Air Blank		0.000	02:57	
Control Test		0.080	02:57	
Air Blank		0.000	02:58	
Subject Sample	# 1	0.178	02:58	
Breath Volume		3.394	02:58	
Air Blank		0.000	02:59	
Air Blank		0.000	03:01	
Subject Sample	# 2	0.174	03:01	
Breath Volume		3.191	03:01	
Air Blank		0.000	03:02	
Control Test		0.074	03:02	Interferent Detect
Air Blank		0.000	03:03	

Data Download Date/Time: 02/01/2018 14:37

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JASIUKIEWICZ,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 01/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 01/14/2018 03:10:33

Diagnostics Check		OK	03:04
Air Blank		0.000	03:04
Control Test		0.079	03:05
Air Blank		0.000	03:05
Subject Sample	# 1	0.166	03:06
Breath Volume		2.664	03:06
Air Blank		0.000	03:06
Air Blank		0.000	03:08
Subject Sample	# 2	0.174	03:08
Breath Volume		3.535	03:08
Air Blank		0.000	03:09
Control Test		0.079	03:09
Air Blank		0.000	03:10
Diagnostics Check		OK	03:10

Data Download Date/Time: 02/01/2018 14:37

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOYLE,THOMAS,E
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: BOYLE,THOMAS
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 01/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: MANN

DATE/TIME OF TEST: 01/25/2018 19:22:08

Diagnostics Check		OK	19:10	
Air Blank		0.000	19:11	
Control Test		0.080	19:11	
Air Blank		0.000	19:12	
Subject Sample	# 1	0.000	19:15	Volume Not Met
Breath Volume		0.277	19:15	
Air Blank		0.000	19:15	
Air Blank		0.000	19:17	
Subject Sample	# 2	0.000	19:20	Volume Not Met
Breath Volume		0.355	19:20	
Air Blank		0.000	19:21	
Control Test		0.080	19:21	
Air Blank		0.000	19:22	
Diagnostics Check		OK	19:22	

Data Download Date/Time: 02/01/2018 14:37

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,JEREMY,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SARVER,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: GROSZ

DATE/TIME OF TEST: 01/01/2018 00:28:35

Diagnostics Check		OK	00:16	
Air Blank		0.000	00:16	
Control Test		0.081	00:16	
Air Blank		0.000	00:17	
Subject Sample	# 1	0.198	00:20	Volume Not Met
Breath Volume		0.523	00:20	
Air Blank		0.000	00:21	
Air Blank		0.000	00:23	
Subject Sample	# 2	0.192	00:23	
Breath Volume		2.039	00:23	
Air Blank		0.000	00:24	
Air Blank		0.000	00:25	
Subject Sample	# 3	0.185	00:27	
Breath Volume		1.550	00:27	
Air Blank		0.000	00:27	
Control Test		0.077	00:28	
Air Blank		0.000	00:28	
Diagnostics Check		OK	00:28	

Data Download Date/Time: 02/12/2018 08:15

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEUTHNER,N,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: LEUTHNER,N
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: HINES

DATE/TIME OF TEST: 01/01/2018 01:48:32

Diagnostics Check		OK	01:36	
Air Blank		0.000	01:36	
Control Test		0.081	01:37	
Air Blank		0.000	01:37	
Subject Sample	# 1	0.156	01:40	Volume Not Met
Breath Volume		0.562	01:40	
Air Blank		0.000	01:41	
Air Blank		0.000	01:43	
Subject Sample	# 2	0.160	01:43	
Breath Volume		1.410	01:43	
Air Blank		0.000	01:44	
Air Blank		0.000	01:46	
Subject Sample	# 3	0.154	01:47	
Breath Volume		1.535	01:47	
Air Blank		0.000	01:47	
Control Test		0.079	01:47	
Air Blank		0.000	01:48	
Diagnostics Check		OK	01:48	

Data Download Date/Time: 02/12/2018 08:15

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLAIN,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SCHULTE,A
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:27

SUBJECT'S LAST NAME: WITT

DATE/TIME OF TEST: 01/01/2018 22:02:55

Diagnostics Check		OK	21:48	
Air Blank		0.000	21:48	
Control Test		0.081	21:48	
Air Blank		0.000	21:49	
Subject Sample	# 1	0.207	21:52	Volume Not Met
Breath Volume		0.945	21:52	
Air Blank		0.000	21:53	
Air Blank		0.000	21:54	
Subject Sample	# 2	0.238	21:55	
Breath Volume		1.558	21:55	
Air Blank		0.000	21:56	
Air Blank		0.000	21:58	
Subject Sample	# 3	0.217	22:01	Volume Not Met
Breath Volume		0.906	22:01	
Air Blank		0.000	22:01	
Control Test		0.079	22:02	
Air Blank		0.000	22:02	
Diagnostics Check		OK	22:02	

Data Download Date/Time: 02/12/2018 08:15

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,JEREMY,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: FRASER,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 16:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 01/04/2018 19:52:19

Diagnostics Check		OK	19:46
Air Blank		0.000	19:46
Control Test		0.081	19:46
Air Blank		0.000	19:47
Subject Sample	# 1	0.000	19:47
Breath Volume		1.675	19:47
Air Blank		0.000	19:48
Air Blank		0.000	19:50
Subject Sample	# 2	0.000	19:50
Breath Volume		1.980	19:50
Air Blank		0.000	19:51
Control Test		0.081	19:51
Air Blank		0.000	19:52
Diagnostics Check		OK	19:52

Data Download Date/Time: 02/12/2018 08:15

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,JEREMY,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MORSE,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 01/05/2018 00:13:58

Diagnostics Check		OK	00:07
Air Blank		0.000	00:08
Control Test		0.081	00:08
Air Blank		0.000	00:08
Subject Sample	# 1	0.174	00:09
Breath Volume		2.753	00:09
Air Blank		0.000	00:10
Air Blank		0.000	00:11
Subject Sample	# 2	0.168	00:12
Breath Volume		2.703	00:12
Air Blank		0.000	00:13
Control Test		0.081	00:13
Air Blank		0.000	00:13
Diagnostics Check		OK	00:13

Data Download Date/Time: 02/12/2018 08:15

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLATZ,KEVIN,L
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: JONES,JASON
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: SENNA

DATE/TIME OF TEST: 01/05/2018 01:16:47

Diagnostics Check		OK	01:08
Air Blank		0.000	01:09
Control Test		0.081	01:09
Air Blank		0.000	01:10
Subject Sample	# 1	0.156	01:11
Breath Volume		1.203	01:11
Air Blank		0.000	01:11
Air Blank		0.000	01:13
Subject Sample	# 2	0.150	01:15
Breath Volume		1.792	01:15
Air Blank		0.000	01:15
Control Test		0.081	01:16
Air Blank		0.000	01:16
Diagnostics Check		OK	01:16

Data Download Date/Time: 02/12/2018 08:15

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGAN,MELISSA,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SILLS,M
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: NAVARRETE

DATE/TIME OF TEST: 01/06/2018 20:33:16

Diagnostics Check		OK	20:26
Air Blank		0.000	20:27
Control Test		0.081	20:27
Air Blank		0.000	20:27
Subject Sample	# 1	0.213	20:28
Breath Volume		2.171	20:28
Air Blank		0.000	20:29
Air Blank		0.000	20:30
Subject Sample	# 2	0.216	20:31
Breath Volume		2.828	20:31
Air Blank		0.000	20:32
Control Test		0.081	20:32
Air Blank		0.000	20:33
Diagnostics Check		OK	20:33

Data Download Date/Time: 02/12/2018 08:15

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLATZ,KEVIN,L
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: JIMENEZ,KELVIN
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: RENDON

DATE/TIME OF TEST: 01/10/2018 01:30:07

Diagnostics Check		OK	01:18	
Air Blank		0.000	01:19	
Control Test		0.080	01:19	
Air Blank		0.000	01:20	
Subject Sample	# 1	0.184	01:23	Volume Not Met
Breath Volume		0.230	01:23	
Air Blank		0.000	01:23	
Air Blank		0.000	01:25	
Subject Sample	# 2		01:28	No Sample Provided
Breath Volume		0.000	01:28	
Air Blank		0.000	01:29	
Control Test		0.080	01:29	
Air Blank		0.000	01:30	
Diagnostics Check		OK	01:30	

Data Download Date/Time: 02/12/2018 08:15

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEONARD,ALEYA,D
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: WRASSE,NICHOLAS
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:37

SUBJECT'S LAST NAME: NASON

DATE/TIME OF TEST: 01/14/2018 06:05:34

Diagnostics Check		OK	05:58
Air Blank		0.000	05:59
Control Test		0.082	05:59
Air Blank		0.000	06:00
Subject Sample	# 1	0.128	06:00
Breath Volume		1.355	06:00
Air Blank		0.000	06:01
Air Blank		0.000	06:03
Subject Sample	# 2	0.127	06:04
Breath Volume		2.242	06:04
Air Blank		0.000	06:04
Control Test		0.081	06:04
Air Blank		0.000	06:05
Diagnostics Check		OK	06:05

Data Download Date/Time: 02/12/2018 08:15

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEONARD,ALEYA,D
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SARVER,SCOTT
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:49

SUBJECT'S LAST NAME: MORSE

DATE/TIME OF TEST: 01/14/2018 21:17:06

Diagnostics Check		OK	21:09
Air Blank		0.000	21:10
Control Test		0.081	21:10
Air Blank		0.000	21:11
Subject Sample	# 1	0.133	21:12
Breath Volume		1.375	21:12
Air Blank		0.000	21:12
Air Blank		0.000	21:14
Subject Sample	# 2	0.133	21:15
Breath Volume		2.171	21:15
Air Blank		0.000	21:16
Control Test		0.081	21:16
Air Blank		0.000	21:17
Diagnostics Check		OK	21:17

Data Download Date/Time: 02/12/2018 08:15

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARIAS,MONICA,S
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: DANNER,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 15:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:10

SUBJECT'S LAST NAME: MARSHALL

DATE/TIME OF TEST: 01/17/2018 16:44:04

Diagnostics Check		OK	16:35
Air Blank		0.000	16:36
Control Test		0.077	16:36
Air Blank		0.000	16:36
Subject Sample	# 1	0.266	16:39
Breath Volume		2.550	16:39
Air Blank		0.000	16:40
Air Blank		0.000	16:42
Subject Sample	# 2	0.265	16:42
Breath Volume		2.292	16:42
Air Blank		0.000	16:43
Control Test		0.081	16:43
Air Blank		0.000	16:43
Diagnostics Check		OK	16:44

Data Download Date/Time: 02/12/2018 08:15

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUCKER,DEPUTY,S
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: STAUB,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 16:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:47

SUBJECT'S LAST NAME: VIQUEIRA

DATE/TIME OF TEST: 01/17/2018 17:18:08

Diagnostics Check		OK	17:12
Air Blank		0.000	17:12
Control Test		0.081	17:12
Air Blank		0.000	17:13
Subject Sample	# 1	0.204	17:13
Breath Volume		3.046	17:13
Air Blank		0.000	17:14
Air Blank		0.000	17:16
Subject Sample	# 2	0.206	17:16
Breath Volume		2.761	17:16
Air Blank		0.000	17:17
Control Test		0.078	17:17
Air Blank		0.000	17:18
Diagnostics Check		OK	17:18

Data Download Date/Time: 02/12/2018 08:15

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEONARD,ALEYA,D
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:58

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 01/18/2018 20:29:43

Diagnostics Check		OK	20:21
Air Blank		0.000	20:22
Control Test		0.081	20:22
Air Blank		0.000	20:22
Subject Sample	# 1	0.237	20:24
Breath Volume		2.117	20:24
Air Blank		0.000	20:25
Air Blank		0.000	20:26
Subject Sample	# 2	0.235	20:28
Breath Volume		2.046	20:28
Air Blank		0.000	20:28
Control Test		0.080	20:29
Air Blank		0.000	20:29
Diagnostics Check		OK	20:29

Data Download Date/Time: 02/12/2018 08:16

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLATZ,KEVIN,L
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: NUGUIT

DATE/TIME OF TEST: 01/19/2018 02:43:15

Diagnostics Check		OK	02:35
Air Blank		0.000	02:36
Control Test		0.081	02:36
Air Blank		0.000	02:37
Subject Sample	# 1	0.252	02:38
Breath Volume		3.152	02:38
Air Blank		0.000	02:39
Air Blank		0.000	02:40
Subject Sample	# 2	0.252	02:41
Breath Volume		2.878	02:41
Air Blank		0.000	02:42
Control Test		0.081	02:42
Air Blank		0.000	02:43
Diagnostics Check		OK	02:43

Data Download Date/Time: 02/12/2018 08:16

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMERON,CELESTE,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: LISTER,R
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 15:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: WETHERINGTON

DATE/TIME OF TEST: 01/19/2018 16:27:40

Diagnostics Check		OK	16:20
Air Blank		0.000	16:21
Control Test		0.081	16:21
Air Blank		0.000	16:21
Subject Sample	# 1	0.010	16:23
Breath Volume		2.046	16:23
Air Blank		0.000	16:23
Air Blank		0.000	16:25
Subject Sample	# 2	0.007	16:26
Breath Volume		2.175	16:26
Air Blank		0.000	16:26
Control Test		0.080	16:27
Air Blank		0.000	16:27
Diagnostics Check		OK	16:27

Data Download Date/Time: 02/12/2018 08:16

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAPMAN,VANESSA,N
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: LEE,ELIJAH
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: UNDERWOOD

DATE/TIME OF TEST: 01/20/2018 22:27:31

Diagnostics Check		OK	22:21
Air Blank		0.000	22:21
Control Test		0.080	22:22
Air Blank		0.000	22:22
Subject Sample	# 1	0.162	22:23
Breath Volume		1.542	22:23
Air Blank		0.000	22:23
Air Blank		0.000	22:25
Subject Sample	# 2	0.150	22:26
Breath Volume		1.171	22:26
Air Blank		0.000	22:26
Control Test		0.080	22:26
Air Blank		0.000	22:27
Diagnostics Check		OK	22:27

Data Download Date/Time: 02/12/2018 08:16

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,CHAD,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NUGENT,DAVID
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: KURJAH

DATE/TIME OF TEST: 01/12/2018 02:27:37

Diagnostics Check		OK	02:20	
Air Blank		0.000	02:20	
Control Test		0.080	02:20	
Air Blank		0.000	02:21	
Subject Sample	# 1		02:24	No Sample Provided
Breath Volume		0.000	02:24	
Air Blank		0.000	02:25	
Air Blank			02:26	RFI Detect
Air Blank			02:27	RFI Detect
Air Blank		0.000	02:27	

Data Download Date/Time: 02/20/2018 19:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,CHAD,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NUGENT,DAVID
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: KURJAH

DATE/TIME OF TEST: 01/12/2018 02:41:39

Diagnostics Check		OK	02:30	
Air Blank		0.000	02:30	
Control Test		0.080	02:30	
Air Blank		0.000	02:31	
Subject Sample	# 1		02:34	No Sample Provided
Breath Volume		0.000	02:34	
Air Blank		0.000	02:35	
Air Blank		0.000	02:37	
Subject Sample	# 2		02:40	No Sample Provided
Breath Volume		0.000	02:40	
Air Blank		0.000	02:40	
Control Test		0.080	02:41	
Air Blank		0.000	02:41	
Diagnostics Check		OK	02:41	

Data Download Date/Time: 02/20/2018 19:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,CHAD,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,DANIEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: LAKY

DATE/TIME OF TEST: 01/12/2018 03:06:22

Diagnostics Check		OK	02:55
Air Blank		0.000	02:55
Control Test		0.080	02:56
Air Blank		0.000	02:56
Subject Sample	# 1	0.131	02:59
Breath Volume		1.285	02:59
Air Blank		0.000	03:00
Air Blank		0.000	03:01
Subject Sample	# 2	0.129	03:04
Breath Volume		1.460	03:04
Air Blank		0.000	03:05
Control Test		0.078	03:05
Air Blank		0.000	03:06
Diagnostics Check		OK	03:06

Data Download Date/Time: 02/20/2018 19:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,CHAD,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: HURT,JONATHAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:43

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 01/12/2018 21:10:34

Diagnostics Check		OK	21:03
Air Blank		0.000	21:04
Control Test		0.080	21:04
Air Blank		0.000	21:04
Subject Sample	# 1	0.019	21:05
Breath Volume		3.593	21:05
Air Blank		0.000	21:06
Air Blank		0.000	21:08
Subject Sample	# 2	0.018	21:09
Breath Volume		2.875	21:09
Air Blank		0.000	21:09
Control Test		0.079	21:09
Air Blank		0.000	21:10
Diagnostics Check		OK	21:10

Data Download Date/Time: 02/20/2018 19:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,CHAD,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NUGENT,DAVID
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: ALLAN

DATE/TIME OF TEST: 01/13/2018 00:32:36

Diagnostics Check		OK	00:25
Air Blank		0.000	00:25
Control Test		0.080	00:26
Air Blank		0.000	00:26
Subject Sample	# 1	0.083	00:27
Breath Volume		1.632	00:27
Air Blank		0.000	00:27
Air Blank		0.000	00:29
Subject Sample	# 2	0.083	00:30
Breath Volume		1.289	00:30
Air Blank		0.000	00:31
Control Test		0.080	00:31
Air Blank		0.000	00:32
Diagnostics Check		OK	00:32

Data Download Date/Time: 02/20/2018 19:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,CHAD,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,DANIEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: WISNIEWSKI

DATE/TIME OF TEST: 01/25/2018 02:25:53

Diagnostics Check		OK	02:18
Air Blank		0.000	02:18
Control Test		0.081	02:19
Air Blank		0.000	02:19
Subject Sample	# 1	0.250	02:21
Breath Volume		3.367	02:21
Air Blank		0.000	02:21
Air Blank		0.000	02:23
Subject Sample	# 2	0.254	02:24
Breath Volume		3.886	02:24
Air Blank		0.000	02:24
Control Test		0.079	02:25
Air Blank		0.000	02:25
Diagnostics Check		OK	02:25

Data Download Date/Time: 02/20/2018 19:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 01/01/2018 02:53:51

Diagnostics Check		OK	02:50	
Air Blank		0.000	02:51	
Control Test		0.082	02:51	
Air Blank		0.000	02:52	
Subject Sample	# 1		02:52	Subject Test Refused
Breath Volume		0.000	02:52	
Air Blank		0.000	02:52	
Control Test		0.082	02:53	
Air Blank		0.000	02:53	
Diagnostics Check		OK	02:53	

Data Download Date/Time: 01/19/2018 14:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GARCIA,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 01/02/2018 01:31:07

Diagnostics Check		OK	01:24
Air Blank		0.000	01:25
Control Test		0.081	01:25
Air Blank		0.000	01:26
Subject Sample	# 1	0.262	01:26
Breath Volume		2.214	01:26
Air Blank		0.000	01:27
Air Blank		0.000	01:29
Subject Sample	# 2	0.270	01:29
Breath Volume		2.414	01:29
Air Blank		0.000	01:30
Control Test		0.081	01:30
Air Blank		0.000	01:31
Diagnostics Check		OK	01:31

Data Download Date/Time: 01/19/2018 14:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BROWN,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: DESSER

DATE/TIME OF TEST: 01/05/2018 03:09:47

Diagnostics Check		OK	03:06	
Air Blank		0.000	03:07	
Control Test		0.083	03:07	
Air Blank		0.000	03:08	
Subject Sample	# 1		03:08	Subject Test Refused
Breath Volume		0.000	03:08	
Air Blank		0.000	03:08	
Control Test		0.083	03:09	
Air Blank		0.000	03:09	
Diagnostics Check		OK	03:09	

Data Download Date/Time: 01/19/2018 14:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARIDA,K,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TORRES,W
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 12/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: PELLICCIOTTI

DATE/TIME OF TEST: 01/05/2018 04:18:49

Diagnostics Check		OK	04:15	
Air Blank		0.000	04:16	
Control Test		0.083	04:16	
Air Blank		0.000	04:17	
Subject Sample	# 1		04:17	Subject Test Refused
Breath Volume		0.000	04:17	
Air Blank		0.000	04:17	
Control Test		0.083	04:18	
Air Blank		0.000	04:18	
Diagnostics Check		OK	04:18	

Data Download Date/Time: 01/19/2018 14:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HOLTON,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:08

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 01/05/2018 05:34:47

Diagnostics Check		OK	05:28
Air Blank		0.000	05:29
Control Test		0.082	05:29
Air Blank		0.000	05:29
Subject Sample	# 1	0.090	05:30
Breath Volume		3.062	05:30
Air Blank		0.000	05:30
Air Blank		0.000	05:32
Subject Sample	# 2	0.083	05:33
Breath Volume		1.785	05:33
Air Blank		0.000	05:33
Control Test		0.081	05:34
Air Blank		0.000	05:34
Diagnostics Check		OK	05:34

Data Download Date/Time: 01/19/2018 14:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: CHI

DATE/TIME OF TEST: 01/06/2018 03:27:39

Diagnostics Check		OK	03:21
Air Blank		0.000	03:21
Control Test		0.082	03:21
Air Blank		0.000	03:22
Subject Sample	# 1	0.083	03:23
Breath Volume		1.464	03:23
Air Blank		0.000	03:23
Air Blank		0.000	03:25
Subject Sample	# 2	0.080	03:26
Breath Volume		1.160	03:26
Air Blank		0.000	03:26
Control Test		0.082	03:27
Air Blank		0.000	03:27
Diagnostics Check		OK	03:27

Data Download Date/Time: 01/19/2018 14:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MULLIS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 01/06/2018 05:43:00

Diagnostics Check		OK	05:40	
Air Blank		0.000	05:40	
Control Test		0.082	05:40	
Air Blank		0.000	05:41	
Subject Sample	# 1		05:41	Subject Test Refused
Breath Volume		0.000	05:41	
Air Blank		0.000	05:42	
Control Test		0.081	05:42	
Air Blank		0.000	05:42	
Diagnostics Check		OK	05:43	

Data Download Date/Time: 01/19/2018 14:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GARCIA,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: HOLMES

DATE/TIME OF TEST: 01/07/2018 00:40:09

Diagnostics Check		OK	00:34
Air Blank		0.000	00:34
Control Test		0.082	00:34
Air Blank		0.000	00:35
Subject Sample	# 1	0.053	00:35
Breath Volume		2.308	00:35
Air Blank		0.000	00:36
Air Blank		0.000	00:38
Subject Sample	# 2	0.055	00:38
Breath Volume		2.109	00:38
Air Blank		0.000	00:39
Control Test		0.082	00:39
Air Blank		0.000	00:40
Diagnostics Check		OK	00:40

Data Download Date/Time: 01/19/2018 14:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 01/07/2018 02:41:37

Diagnostics Check		OK	02:38	
Air Blank		0.000	02:39	
Control Test		0.082	02:39	
Air Blank		0.000	02:40	
Subject Sample	# 1		02:40	Subject Test Refused
Breath Volume		0.000	02:40	
Air Blank		0.000	02:40	
Control Test		0.082	02:41	
Air Blank		0.000	02:41	
Diagnostics Check		OK	02:41	

Data Download Date/Time: 01/19/2018 14:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FILBERT,E
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 12/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 01/07/2018 05:20:53

Diagnostics Check		OK	05:13
Air Blank		0.000	05:14
Control Test		0.083	05:14
Air Blank		0.000	05:15
Subject Sample	# 1	0.106	05:16
Breath Volume		1.683	05:16
Air Blank		0.000	05:16
Air Blank		0.000	05:18
Subject Sample	# 2	0.103	05:19
Breath Volume		1.757	05:19
Air Blank		0.000	05:19
Control Test		0.082	05:20
Air Blank		0.000	05:20
Diagnostics Check		OK	05:20

Data Download Date/Time: 01/19/2018 14:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: SHANNON

DATE/TIME OF TEST: 01/07/2018 23:03:23

Diagnostics Check		OK	23:00	
Air Blank		0.000	23:01	
Control Test		0.081	23:01	
Air Blank		0.000	23:01	
Subject Sample	# 1		23:02	Subject Test Refused
Breath Volume		0.000	23:02	
Air Blank		0.000	23:02	
Control Test		0.081	23:02	
Air Blank		0.000	23:03	
Diagnostics Check		OK	23:03	

Data Download Date/Time: 01/19/2018 14:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: YAKLIN

DATE/TIME OF TEST: 01/08/2018 00:51:02

Diagnostics Check		OK	00:48	
Air Blank		0.000	00:48	
Control Test		0.081	00:48	
Air Blank		0.000	00:49	
Subject Sample	# 1		00:49	Subject Test Refused
Breath Volume		0.000	00:49	
Air Blank		0.000	00:50	
Control Test		0.081	00:50	
Air Blank		0.000	00:50	
Diagnostics Check		OK	00:51	

Data Download Date/Time: 01/19/2018 14:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: NORRIS

DATE/TIME OF TEST: 01/08/2018 02:10:58

Diagnostics Check		OK	02:04
Air Blank		0.000	02:05
Control Test		0.081	02:05
Air Blank		0.000	02:06
Subject Sample	# 1	0.153	02:06
Breath Volume		3.085	02:06
Air Blank		0.000	02:07
Air Blank		0.000	02:09
Subject Sample	# 2	0.155	02:09
Breath Volume		3.019	02:09
Air Blank		0.000	02:10
Control Test		0.081	02:10
Air Blank		0.000	02:10
Diagnostics Check		OK	02:10

Data Download Date/Time: 01/19/2018 14:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: HINES

DATE/TIME OF TEST: 01/08/2018 03:31:27

Diagnostics Check		OK	03:28	
Air Blank		0.000	03:29	
Control Test		0.082	03:29	
Air Blank		0.000	03:29	
Subject Sample	# 1		03:30	Subject Test Refused
Breath Volume		0.000	03:30	
Air Blank		0.000	03:30	
Control Test		0.082	03:30	
Air Blank		0.000	03:31	
Diagnostics Check		OK	03:31	

Data Download Date/Time: 01/19/2018 14:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: PATTERSON

DATE/TIME OF TEST: 01/08/2018 23:05:50

Diagnostics Check		OK	22:59
Air Blank		0.000	23:00
Control Test		0.081	23:00
Air Blank		0.000	23:01
Subject Sample	# 1	0.162	23:01
Breath Volume		2.125	23:01
Air Blank		0.000	23:02
Air Blank		0.000	23:03
Subject Sample	# 2	0.151	23:04
Breath Volume		1.304	23:04
Air Blank		0.000	23:04
Control Test		0.081	23:05
Air Blank		0.000	23:05
Diagnostics Check		OK	23:05

Data Download Date/Time: 01/19/2018 14:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: LUCKEY

DATE/TIME OF TEST: 01/09/2018 02:49:01

Diagnostics Check		OK	02:43
Air Blank		0.000	02:43
Control Test		0.081	02:43
Air Blank		0.000	02:44
Subject Sample	# 1	0.168	02:44
Breath Volume		3.261	02:44
Air Blank		0.000	02:45
Air Blank		0.000	02:47
Subject Sample	# 2	0.167	02:47
Breath Volume		3.578	02:47
Air Blank		0.000	02:48
Control Test		0.081	02:48
Air Blank		0.000	02:48
Diagnostics Check		OK	02:49

Data Download Date/Time: 01/19/2018 14:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AUGUSTINE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: PORTILLO HUESO

DATE/TIME OF TEST: 01/14/2018 02:03:17

Diagnostics Check		OK	02:00	
Air Blank		0.000	02:00	
Control Test		0.081	02:01	
Air Blank		0.000	02:01	
Subject Sample	# 1		02:01	Subject Test Refused
Breath Volume		0.000	02:01	
Air Blank		0.000	02:02	
Control Test		0.081	02:02	
Air Blank		0.000	02:03	
Diagnostics Check		OK	02:03	

Data Download Date/Time: 01/19/2018 14:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: SMOAK

DATE/TIME OF TEST: 01/19/2018 00:57:14

Diagnostics Check		OK	00:54	
Air Blank		0.000	00:54	
Control Test		0.082	00:55	
Air Blank		0.000	00:55	
Subject Sample	# 1		00:55	Subject Test Refused
Breath Volume		0.000	00:55	
Air Blank		0.000	00:56	
Control Test		0.083	00:56	
Air Blank		0.000	00:57	
Diagnostics Check		OK	00:57	

Data Download Date/Time: 01/19/2018 14:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 01/20/2018 00:51:29

Diagnostics Check		OK	00:44
Air Blank		0.000	00:44
Control Test		0.082	00:45
Air Blank		0.000	00:45
Subject Sample	# 1	0.178	00:46
Breath Volume		1.671	00:46
Air Blank		0.000	00:47
Air Blank		0.000	00:49
Subject Sample	# 2	0.192	00:49
Breath Volume		1.449	00:49
Air Blank		0.000	00:50
Control Test		0.082	00:50
Air Blank		0.000	00:51
Diagnostics Check		OK	00:51

Data Download Date/Time: 02/01/2018 13:45

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:02

SUBJECT'S LAST NAME: HEMPHILL

DATE/TIME OF TEST: 01/20/2018 05:29:48

Diagnostics Check		OK	05:23
Air Blank		0.000	05:23
Control Test		0.082	05:23
Air Blank		0.000	05:24
Subject Sample	# 1	0.188	05:25
Breath Volume		2.750	05:25
Air Blank		0.000	05:25
Air Blank		0.000	05:27
Subject Sample	# 2	0.190	05:28
Breath Volume		3.066	05:28
Air Blank		0.000	05:28
Control Test		0.082	05:29
Air Blank		0.000	05:29
Diagnostics Check		OK	05:29

Data Download Date/Time: 02/01/2018 13:45

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRYANT,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: SIMS

DATE/TIME OF TEST: 01/21/2018 01:03:53

Diagnostics Check		OK	01:01	
Air Blank		0.000	01:01	
Control Test		0.081	01:01	
Air Blank		0.000	01:02	
Subject Sample	# 1		01:02	Subject Test Refused
Breath Volume		0.000	01:02	
Air Blank		0.000	01:02	
Control Test		0.082	01:03	
Air Blank		0.000	01:03	
Diagnostics Check		OK	01:03	

Data Download Date/Time: 02/01/2018 13:45

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DRAGOMIR-RAMIREZ,IOANA
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: GEARHART

DATE/TIME OF TEST: 01/21/2018 02:41:34

Diagnostics Check		OK	02:33
Air Blank		0.000	02:33
Control Test		0.082	02:34
Air Blank		0.000	02:34
Subject Sample	# 1	0.211	02:36
Breath Volume		3.406	02:36
Air Blank		0.000	02:36
Air Blank		0.000	02:38
Subject Sample	# 2	0.225	02:40
Breath Volume		1.625	02:40
Air Blank		0.000	02:40
Control Test		0.081	02:40
Air Blank		0.000	02:41
Diagnostics Check		OK	02:41

Data Download Date/Time: 02/01/2018 13:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: WELCH

DATE/TIME OF TEST: 01/21/2018 03:55:37

Diagnostics Check		OK	03:47
Air Blank		0.000	03:48
Control Test		0.082	03:48
Air Blank		0.000	03:49
Subject Sample	# 1	0.127	03:50
Breath Volume		1.328	03:50
Air Blank		0.000	03:51
Air Blank		0.000	03:53
Subject Sample	# 2	0.139	03:54
Breath Volume		1.281	03:54
Air Blank		0.000	03:54
Control Test		0.081	03:54
Air Blank		0.000	03:55
Diagnostics Check		OK	03:55

Data Download Date/Time: 02/01/2018 13:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SEEGOBIN,V
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: MUNOZ RAZO

DATE/TIME OF TEST: 01/21/2018 05:42:28

Diagnostics Check		OK	05:35
Air Blank		0.000	05:36
Control Test		0.081	05:36
Air Blank		0.000	05:37
Subject Sample	# 1	0.161	05:37
Breath Volume		1.332	05:37
Air Blank		0.000	05:38
Air Blank		0.000	05:39
Subject Sample	# 2	0.174	05:40
Breath Volume		1.414	05:40
Air Blank		0.000	05:41
Control Test		0.081	05:41
Air Blank		0.000	05:42
Diagnostics Check		OK	05:42

Data Download Date/Time: 02/01/2018 13:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 01/21/2018 23:56:35

Diagnostics Check		OK	23:50
Air Blank		0.000	23:50
Control Test		0.081	23:51
Air Blank		0.000	23:51
Subject Sample	# 1	0.144	23:52
Breath Volume		3.812	23:52
Air Blank		0.000	23:52
Air Blank		0.000	23:54
Subject Sample	# 2	0.142	23:55
Breath Volume		3.828	23:55
Air Blank		0.000	23:55
Control Test		0.081	23:55
Air Blank		0.000	23:56
Diagnostics Check		OK	23:56

Data Download Date/Time: 02/01/2018 13:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 01/22/2018 01:55:32

Diagnostics Check		OK	01:49
Air Blank		0.000	01:49
Control Test		0.081	01:49
Air Blank		0.000	01:50
Subject Sample	# 1	0.207	01:50
Breath Volume		1.453	01:50
Air Blank		0.000	01:51
Air Blank		0.000	01:53
Subject Sample	# 2	0.213	01:53
Breath Volume		2.562	01:53
Air Blank		0.000	01:54
Control Test		0.081	01:54
Air Blank		0.000	01:55
Diagnostics Check		OK	01:55

Data Download Date/Time: 02/01/2018 13:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMONS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: WHIPPLE

DATE/TIME OF TEST: 01/22/2018 04:57:39

Diagnostics Check		OK	04:51
Air Blank		0.000	04:51
Control Test		0.082	04:52
Air Blank		0.000	04:52
Subject Sample	# 1	0.113	04:53
Breath Volume		1.492	04:53
Air Blank		0.000	04:53
Air Blank		0.000	04:55
Subject Sample	# 2	0.115	04:56
Breath Volume		1.597	04:56
Air Blank		0.000	04:56
Control Test		0.082	04:57
Air Blank		0.000	04:57
Diagnostics Check		OK	04:57

Data Download Date/Time: 02/01/2018 13:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 01/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: POWELL

DATE/TIME OF TEST: 01/23/2018 23:03:19

Diagnostics Check		OK	23:00	
Air Blank		0.000	23:00	
Control Test		0.081	23:01	
Air Blank		0.000	23:01	
Subject Sample	# 1		23:01	Subject Test Refused
Breath Volume		0.000	23:01	
Air Blank		0.000	23:02	
Control Test		0.081	23:02	
Air Blank		0.000	23:03	
Diagnostics Check		OK	23:03	

Data Download Date/Time: 02/01/2018 13:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MARRERO,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: SHABAZZ

DATE/TIME OF TEST: 01/24/2018 01:24:59

Diagnostics Check		OK	01:19	
Air Blank		0.000	01:20	
Control Test		0.081	01:20	
Air Blank		0.000	01:21	
Subject Sample	# 1		01:23	Subject Test Refused
Breath Volume		1.308	01:23	
Air Blank		0.000	01:24	
Control Test		0.082	01:24	
Air Blank		0.000	01:24	
Diagnostics Check		OK	01:24	

Data Download Date/Time: 02/01/2018 13:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 01/24/2018 04:09:16

Diagnostics Check		OK	04:03
Air Blank		0.000	04:03
Control Test		0.082	04:04
Air Blank		0.000	04:04
Subject Sample	# 1	0.149	04:04
Breath Volume		2.902	04:04
Air Blank		0.000	04:05
Air Blank		0.000	04:07
Subject Sample	# 2	0.152	04:07
Breath Volume		2.945	04:07
Air Blank		0.000	04:08
Control Test		0.082	04:08
Air Blank		0.000	04:09
Diagnostics Check		OK	04:09

Data Download Date/Time: 02/01/2018 13:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BROWN,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: ARAYA FERNANDEZ

DATE/TIME OF TEST: 01/25/2018 00:43:04

Diagnostics Check		OK	00:36
Air Blank		0.000	00:36
Control Test		0.081	00:36
Air Blank		0.000	00:37
Subject Sample	# 1	0.250	00:38
Breath Volume		1.789	00:38
Air Blank		0.000	00:38
Air Blank		0.000	00:40
Subject Sample	# 2	0.255	00:41
Breath Volume		1.828	00:41
Air Blank		0.000	00:42
Control Test		0.082	00:42
Air Blank		0.000	00:42
Diagnostics Check		OK	00:43

Data Download Date/Time: 02/01/2018 13:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: CRUMRINE

DATE/TIME OF TEST: 01/25/2018 02:34:47

Diagnostics Check		OK	02:28
Air Blank		0.000	02:29
Control Test		0.081	02:29
Air Blank		0.000	02:30
Subject Sample	# 1	0.150	02:30
Breath Volume		1.746	02:30
Air Blank		0.000	02:31
Air Blank		0.000	02:32
Subject Sample	# 2	0.140	02:33
Breath Volume		1.527	02:33
Air Blank		0.000	02:33
Control Test		0.081	02:34
Air Blank		0.000	02:34
Diagnostics Check		OK	02:34

Data Download Date/Time: 02/01/2018 13:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: GLENN

DATE/TIME OF TEST: 01/25/2018 03:53:26

Diagnostics Check		OK	03:50	
Air Blank		0.000	03:51	
Control Test		0.081	03:51	
Air Blank		0.000	03:51	
Subject Sample	# 1		03:52	Subject Test Refused
Breath Volume		0.000	03:52	
Air Blank		0.000	03:52	
Control Test		0.082	03:52	
Air Blank		0.000	03:53	
Diagnostics Check		OK	03:53	

Data Download Date/Time: 02/01/2018 13:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: BOWMAN

DATE/TIME OF TEST: 01/25/2018 22:56:30

Diagnostics Check		OK	22:53	
Air Blank		0.000	22:53	
Control Test		0.082	22:54	
Air Blank		0.000	22:54	
Subject Sample	# 1		22:55	Subject Test Refused
Breath Volume		0.000	22:55	
Air Blank		0.000	22:55	
Control Test		0.082	22:55	
Air Blank		0.000	22:56	
Diagnostics Check		OK	22:56	

Data Download Date/Time: 02/01/2018 13:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 01/26/2018 04:10:12

Diagnostics Check		OK	04:03
Air Blank		0.000	04:03
Control Test		0.081	04:04
Air Blank		0.000	04:04
Subject Sample	# 1	0.124	04:05
Breath Volume		2.617	04:05
Air Blank		0.000	04:06
Air Blank		0.000	04:08
Subject Sample	# 2	0.131	04:08
Breath Volume		1.351	04:08
Air Blank		0.000	04:09
Control Test		0.081	04:09
Air Blank		0.000	04:10
Diagnostics Check		OK	04:10

Data Download Date/Time: 02/01/2018 13:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,ROBERTS,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: COUGHLIN

DATE/TIME OF TEST: 01/26/2018 04:57:28

Diagnostics Check		OK	04:51
Air Blank		0.000	04:51
Control Test		0.081	04:52
Air Blank		0.000	04:52
Subject Sample	# 1	0.111	04:53
Breath Volume		2.011	04:53
Air Blank		0.000	04:53
Air Blank		0.000	04:55
Subject Sample	# 2	0.111	04:55
Breath Volume		1.800	04:55
Air Blank		0.000	04:56
Control Test		0.082	04:56
Air Blank		0.000	04:57
Diagnostics Check		OK	04:57

Data Download Date/Time: 02/01/2018 13:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: GRAHAM

DATE/TIME OF TEST: 01/27/2018 01:43:09

Diagnostics Check		OK	01:36
Air Blank		0.000	01:37
Control Test		0.081	01:37
Air Blank		0.000	01:38
Subject Sample	# 1	0.191	01:38
Breath Volume		3.140	01:38
Air Blank		0.000	01:39
Air Blank		0.000	01:41
Subject Sample	# 2	0.181	01:41
Breath Volume		2.734	01:41
Air Blank		0.000	01:42
Control Test		0.081	01:42
Air Blank		0.000	01:43
Diagnostics Check		OK	01:43

Data Download Date/Time: 02/01/2018 13:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: GRAY

DATE/TIME OF TEST: 01/27/2018 02:32:48

Diagnostics Check		OK	02:27
Air Blank		0.000	02:27
Control Test		0.081	02:27
Air Blank		0.000	02:28
Subject Sample	# 1	0.000	02:28
Breath Volume		2.843	02:28
Air Blank		0.000	02:29
Air Blank		0.000	02:31
Subject Sample	# 2	0.000	02:31
Breath Volume		2.101	02:31
Air Blank		0.000	02:31
Control Test		0.081	02:32
Air Blank		0.000	02:32
Diagnostics Check		OK	02:32

Data Download Date/Time: 02/01/2018 13:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FILBERT,E
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 01/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: SWART

DATE/TIME OF TEST: 01/27/2018 05:01:32

Diagnostics Check		OK	04:55
Air Blank		0.000	04:55
Control Test		0.081	04:56
Air Blank		0.000	04:56
Subject Sample	# 1	0.170	04:56
Breath Volume		1.519	04:56
Air Blank		0.000	04:57
Air Blank		0.000	04:59
Subject Sample	# 2	0.172	05:00
Breath Volume		1.320	05:00
Air Blank		0.000	05:00
Control Test		0.082	05:00
Air Blank		0.000	05:01
Diagnostics Check		OK	05:01

Data Download Date/Time: 02/01/2018 13:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:27

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 01/27/2018 05:54:05

Diagnostics Check		OK	05:47
Air Blank		0.000	05:48
Control Test		0.080	05:48
Air Blank		0.000	05:49
Subject Sample	# 1	0.197	05:49
Breath Volume		1.316	05:49
Air Blank		0.000	05:50
Air Blank		0.000	05:51
Subject Sample	# 2	0.200	05:52
Breath Volume		2.253	05:52
Air Blank		0.000	05:53
Control Test		0.081	05:53
Air Blank		0.000	05:54
Diagnostics Check		OK	05:54

Data Download Date/Time: 02/01/2018 13:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BINKLEY,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 01/28/2018 02:33:23

Diagnostics Check		OK	02:27
Air Blank		0.000	02:27
Control Test		0.080	02:28
Air Blank		0.000	02:28
Subject Sample	# 1	0.097	02:28
Breath Volume		1.371	02:28
Air Blank		0.000	02:29
Air Blank		0.000	02:31
Subject Sample	# 2	0.090	02:31
Breath Volume		1.296	02:31
Air Blank		0.000	02:32
Control Test		0.081	02:32
Air Blank		0.000	02:33
Diagnostics Check		OK	02:33

Data Download Date/Time: 02/01/2018 13:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: COUCH,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: ABRAHAM

DATE/TIME OF TEST: 01/28/2018 03:45:49

Diagnostics Check		OK	03:39
Air Blank		0.000	03:40
Control Test		0.081	03:40
Air Blank		0.000	03:41
Subject Sample	# 1	0.145	03:41
Breath Volume		1.632	03:41
Air Blank		0.000	03:42
Air Blank		0.000	03:43
Subject Sample	# 2	0.148	03:44
Breath Volume		2.113	03:44
Air Blank		0.000	03:44
Control Test		0.081	03:45
Air Blank		0.000	03:45
Diagnostics Check		OK	03:45

Data Download Date/Time: 02/01/2018 13:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 01/28/2018 04:26:29

Diagnostics Check		OK	04:20
Air Blank		0.000	04:20
Control Test		0.081	04:21
Air Blank		0.000	04:21
Subject Sample	# 1	0.103	04:22
Breath Volume		4.160	04:22
Air Blank		0.000	04:22
Air Blank		0.000	04:24
Subject Sample	# 2	0.108	04:24
Breath Volume		4.511	04:24
Air Blank		0.000	04:25
Control Test		0.081	04:25
Air Blank		0.000	04:26
Diagnostics Check		OK	04:26

Data Download Date/Time: 02/01/2018 13:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEE,JAMAL,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HAWKINS,A
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 01/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:49

SUBJECT'S LAST NAME: WALTERS

DATE/TIME OF TEST: 01/28/2018 08:14:58

Diagnostics Check		OK	08:12	
Air Blank		0.000	08:12	
Control Test		0.081	08:12	
Air Blank		0.000	08:13	
Subject Sample	# 1		08:13	Subject Test Refused
Breath Volume		0.000	08:13	
Air Blank		0.000	08:14	
Control Test		0.081	08:14	
Air Blank		0.000	08:14	
Diagnostics Check		OK	08:14	

Data Download Date/Time: 02/01/2018 13:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:46

SUBJECT'S LAST NAME: STANLEY

DATE/TIME OF TEST: 01/28/2018 23:09:40

Diagnostics Check		OK	23:06	
Air Blank		0.000	23:07	
Control Test		0.080	23:07	
Air Blank		0.000	23:08	
Subject Sample	# 1		23:08	Subject Test Refused
Breath Volume		0.000	23:08	
Air Blank		0.000	23:08	
Control Test		0.080	23:09	
Air Blank		0.000	23:09	
Diagnostics Check		OK	23:09	

Data Download Date/Time: 02/01/2018 13:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 01/29/2018 00:01:14

Diagnostics Check		OK	23:55
Air Blank		0.000	23:55
Control Test		0.080	23:55
Air Blank		0.000	23:56
Subject Sample	# 1	0.260	23:56
Breath Volume		1.738	23:56
Air Blank		0.000	23:57
Air Blank		0.000	23:59
Subject Sample	# 2	0.275	23:59
Breath Volume		1.347	23:59
Air Blank		0.000	00:00
Control Test		0.081	00:00
Air Blank		0.000	00:01
Diagnostics Check		OK	00:01

Data Download Date/Time: 02/01/2018 13:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GARCIA,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: CONTRERAS ANDRADE

DATE/TIME OF TEST: 01/29/2018 02:30:16

Diagnostics Check		OK	02:27	
Air Blank		0.000	02:27	
Control Test		0.081	02:28	
Air Blank		0.000	02:28	
Subject Sample	# 1		02:28	Subject Test Refused
Breath Volume		0.000	02:28	
Air Blank		0.000	02:29	
Control Test		0.080	02:29	
Air Blank		0.000	02:30	
Diagnostics Check		OK	02:30	

Data Download Date/Time: 02/01/2018 13:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,D,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MATTHEWS,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 10:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:41

SUBJECT'S LAST NAME: QUIJANO

DATE/TIME OF TEST: 01/29/2018 12:29:34

Diagnostics Check		OK	12:26	
Air Blank		0.000	12:27	
Control Test		0.081	12:27	
Air Blank		0.000	12:27	
Subject Sample	# 1		12:28	Subject Test Refused
Breath Volume		0.000	12:28	
Air Blank		0.000	12:28	
Control Test		0.081	12:28	
Air Blank		0.000	12:29	
Diagnostics Check		OK	12:29	

Data Download Date/Time: 02/01/2018 13:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HOOVER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: BATES

DATE/TIME OF TEST: 01/29/2018 23:57:45

Diagnostics Check		OK	23:54	
Air Blank		0.000	23:55	
Control Test		0.081	23:55	
Air Blank		0.000	23:56	
Subject Sample	# 1		23:56	Subject Test Refused
Breath Volume		0.000	23:56	
Air Blank		0.000	23:56	
Control Test		0.081	23:57	
Air Blank		0.000	23:57	
Diagnostics Check		OK	23:57	

Data Download Date/Time: 02/01/2018 13:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 01/30/2018 00:44:57

Diagnostics Check		OK	00:42	
Air Blank		0.000	00:42	
Control Test		0.081	00:42	
Air Blank		0.000	00:43	
Subject Sample	# 1		00:43	Subject Test Refused
Breath Volume		0.000	00:43	
Air Blank		0.000	00:44	
Control Test		0.082	00:44	
Air Blank		0.000	00:44	
Diagnostics Check		OK	00:44	

Data Download Date/Time: 02/01/2018 13:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 01/30/2018 04:27:20

Diagnostics Check		OK	04:21
Air Blank		0.000	04:21
Control Test		0.081	04:22
Air Blank		0.000	04:22
Subject Sample	# 1	0.196	04:23
Breath Volume		1.890	04:23
Air Blank		0.000	04:23
Air Blank		0.000	04:25
Subject Sample	# 2	0.202	04:25
Breath Volume		2.500	04:25
Air Blank		0.000	04:26
Control Test		0.081	04:26
Air Blank		0.000	04:27
Diagnostics Check		OK	04:27

Data Download Date/Time: 02/01/2018 13:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALAGON,ODILA,I
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: PAIGE,LEON
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 12/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 03:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: MALAGON

DATE/TIME OF TEST: 01/05/2018 04:33:23

Diagnostics Check		OK	04:27
Air Blank		0.000	04:27
Control Test		0.081	04:28
Air Blank		0.000	04:28
Subject Sample	# 1	0.209	04:29
Breath Volume		1.953	04:29
Air Blank		0.000	04:29
Air Blank		0.000	04:31
Subject Sample	# 2	0.197	04:31
Breath Volume		1.460	04:31
Air Blank		0.000	04:32
Control Test		0.080	04:32
Air Blank		0.000	04:33
Diagnostics Check		OK	04:33

Data Download Date/Time: 01/31/2018 10:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,GIRANNY,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: GAVALLAS,ELIZABETH
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 12/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 17:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:23

SUBJECT'S LAST NAME: ROMERO

DATE/TIME OF TEST: 01/07/2018 18:39:04

Diagnostics Check		OK	18:32
Air Blank		0.000	18:33
Control Test		0.080	18:33
Air Blank		0.000	18:33
Subject Sample	# 1	0.162	18:34
Breath Volume		1.949	18:34
Air Blank		0.000	18:35
Air Blank		0.000	18:36
Subject Sample	# 2	0.168	18:37
Breath Volume		1.523	18:37
Air Blank		0.000	18:38
Control Test		0.080	18:38
Air Blank		0.000	18:38
Diagnostics Check		OK	18:39

Data Download Date/Time: 01/31/2018 10:35

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUMPHRY,KERRY,R
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: CASTILLO,RICARDO
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 12/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: GINZBURG

DATE/TIME OF TEST: 01/09/2018 23:26:28

Diagnostics Check		OK	23:23	
Air Blank		0.000	23:24	
Control Test		0.079	23:24	
Air Blank		0.000	23:24	
Subject Sample	# 1		23:25	Subject Test Refused
Breath Volume		0.000	23:25	
Air Blank		0.000	23:25	
Control Test		0.079	23:25	
Air Blank		0.000	23:26	
Diagnostics Check		OK	23:26	

Data Download Date/Time: 01/31/2018 10:35

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORTON,REGINALD,B
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: CALLAN,ANTHONY
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 12/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: NARANJO

DATE/TIME OF TEST: 01/10/2018 01:29:35

Diagnostics Check		OK	01:23
Air Blank		0.000	01:23
Control Test		0.079	01:24
Air Blank		0.000	01:24
Subject Sample	# 1	0.245	01:25
Breath Volume		1.578	01:25
Air Blank		0.000	01:25
Air Blank		0.000	01:27
Subject Sample	# 2	0.245	01:28
Breath Volume		2.195	01:28
Air Blank		0.000	01:28
Control Test		0.079	01:29
Air Blank		0.000	01:29
Diagnostics Check		OK	01:29

Data Download Date/Time: 01/31/2018 10:35

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CREWS,CURTIS,L
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: PAIGE,LEON
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 12/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: ALMY

DATE/TIME OF TEST: 01/13/2018 01:21:13

Diagnostics Check		OK	01:16	
Air Blank		0.000	01:16	
Control Test		0.079	01:16	
Air Blank		0.000	01:17	
Subject Sample	# 1		01:19	Subject Test Refused
Breath Volume		0.000	01:19	
Air Blank		0.000	01:20	
Control Test		0.080	01:20	
Air Blank		0.000	01:21	
Diagnostics Check		OK	01:21	

Data Download Date/Time: 01/31/2018 10:35

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONFESSORE,STEVE,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: PAIGE,LEON
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 12/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: WONGWAI

DATE/TIME OF TEST: 01/14/2018 04:07:56

Diagnostics Check		OK	04:00
Air Blank		0.000	04:01
Control Test		0.080	04:01
Air Blank		0.000	04:02
Subject Sample	# 1	0.148	04:03
Breath Volume		2.285	04:03
Air Blank		0.000	04:03
Air Blank		0.000	04:05
Subject Sample	# 2	0.148	04:06
Breath Volume		2.558	04:06
Air Blank		0.000	04:07
Control Test		0.079	04:07
Air Blank		0.000	04:07
Diagnostics Check		OK	04:07

Data Download Date/Time: 01/31/2018 10:35

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,JOSE,D
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: PEREZ,JOSE
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 12/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 17:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:06

SUBJECT'S LAST NAME: VALLE

DATE/TIME OF TEST: 01/14/2018 18:35:17

Diagnostics Check		OK	18:29
Air Blank		0.000	18:29
Control Test		0.081	18:29
Air Blank		0.000	18:30
Subject Sample	# 1	0.226	18:30
Breath Volume		1.781	18:30
Air Blank		0.000	18:31
Air Blank		0.000	18:33
Subject Sample	# 2	0.232	18:33
Breath Volume		3.261	18:33
Air Blank		0.000	18:34
Control Test		0.079	18:34
Air Blank		0.000	18:35
Diagnostics Check		OK	18:35

Data Download Date/Time: 01/31/2018 10:35

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CREWS,CURTIS,L
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: CASTILLO,RICARDO
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 12/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: ARRECHEA

DATE/TIME OF TEST: 01/14/2018 23:33:06

Diagnostics Check		OK	23:25
Air Blank		0.000	23:26
Control Test		0.079	23:26
Air Blank		0.000	23:26
Subject Sample	# 1	0.133	23:28
Breath Volume		1.222	23:28
Air Blank		0.000	23:28
Air Blank		0.000	23:30
Subject Sample	# 2	0.127	23:31
Breath Volume		1.246	23:31
Air Blank		0.000	23:32
Control Test		0.079	23:32
Air Blank		0.000	23:33
Diagnostics Check		OK	23:33

Data Download Date/Time: 01/31/2018 10:35

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONFESSORE,STEVE,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: PAIGE,LEON
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 12/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 01/20/2018 01:16:28

Diagnostics Check		OK	01:13	
Air Blank		0.000	01:13	
Control Test		0.079	01:14	
Air Blank		0.000	01:14	
Subject Sample	# 1		01:15	Subject Test Refused
Breath Volume		0.000	01:15	
Air Blank		0.000	01:15	
Control Test		0.081	01:15	
Air Blank		0.000	01:16	
Diagnostics Check		OK	01:16	

Data Download Date/Time: 01/31/2018 10:35

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONFESSORE,STEVE,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: PAIGE,LEON
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 12/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 05:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: SILVERA

DATE/TIME OF TEST: 01/20/2018 05:52:58

Diagnostics Check		OK	05:45
Air Blank		0.000	05:46
Control Test		0.080	05:46
Air Blank		0.000	05:47
Subject Sample	# 1	0.139	05:48
Breath Volume		4.167	05:48
Air Blank		0.000	05:49
Air Blank		0.000	05:50
Subject Sample	# 2	0.141	05:51
Breath Volume		4.089	05:51
Air Blank		0.000	05:52
Control Test		0.079	05:52
Air Blank		0.000	05:52
Diagnostics Check		OK	05:52

Data Download Date/Time: 01/31/2018 10:35

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CREWS,CURTIS,L
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: PAIGE,LEON
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 12/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: VASQUEZ

DATE/TIME OF TEST: 01/21/2018 02:09:16

Diagnostics Check		OK	02:06	
Air Blank		0.000	02:06	
Control Test		0.080	02:07	
Air Blank		0.000	02:07	
Subject Sample	# 1		02:07	Subject Test Refused
Breath Volume		0.000	02:07	
Air Blank		0.000	02:08	
Control Test		0.079	02:08	
Air Blank		0.000	02:09	
Diagnostics Check		OK	02:09	

Data Download Date/Time: 01/31/2018 10:35

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAIGE,LEON,I
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: CHEVALIER,ISMAEL
ARRESTING OFFICER AGENCY: North Bay Village PD
LAST AGENCY INSPECTION: 12/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 22:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: BENITES

DATE/TIME OF TEST: 01/21/2018 23:16:09

Diagnostics Check		OK	23:13	
Air Blank		0.000	23:13	
Control Test		0.080	23:14	
Air Blank		0.000	23:14	
Subject Sample	# 1		23:14	Subject Test Refused
Breath Volume		0.000	23:14	
Air Blank		0.000	23:15	
Control Test		0.079	23:15	
Air Blank		0.000	23:16	
Diagnostics Check		OK	23:16	

Data Download Date/Time: 01/31/2018 10:35

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONSALVO,ESTEBAN,A
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: MUI,KINSHUN
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 12/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 01/14/2018 01:44:42

Diagnostics Check		OK	01:41	
Air Blank		0.000	01:41	
Control Test		0.079	01:42	
Air Blank		0.000	01:42	
Subject Sample	# 1		01:43	Subject Test Refused
Breath Volume		0.000	01:43	
Air Blank		0.000	01:43	
Control Test		0.079	01:44	
Air Blank		0.000	01:44	
Diagnostics Check		OK	01:44	

Data Download Date/Time: 01/24/2018 12:16

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,G,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,G
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 12/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:46

SUBJECT'S LAST NAME: GUGLIELMO

DATE/TIME OF TEST: 01/10/2018 22:18:49

Diagnostics Check		OK	22:12
Air Blank		0.000	22:12
Control Test		0.079	22:12
Air Blank		0.000	22:13
Subject Sample	# 1	0.238	22:13
Breath Volume		1.500	22:13
Air Blank		0.000	22:14
Air Blank		0.000	22:16
Subject Sample	# 2	0.228	22:17
Breath Volume		1.511	22:17
Air Blank		0.000	22:17
Control Test		0.079	22:18
Air Blank		0.000	22:18
Diagnostics Check		OK	22:18

Data Download Date/Time: 01/30/2018 10:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,G,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,G
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 12/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: PETTY

DATE/TIME OF TEST: 01/16/2018 22:06:34

Diagnostics Check		OK	21:57	
Air Blank		0.000	21:58	
Control Test		0.080	21:58	
Air Blank		0.000	21:59	
Subject Sample	# 1	0.124	22:02	Volume Not Met
Breath Volume		0.785	22:02	
Air Blank		0.000	22:03	
Air Blank		0.000	22:04	
Subject Sample	# 2		22:05	Subject Test Refused
Breath Volume		0.000	22:05	
Air Blank		0.000	22:05	
Control Test		0.076	22:05	
Air Blank		0.000	22:06	
Diagnostics Check		OK	22:06	

Data Download Date/Time: 01/30/2018 10:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,JEFFREY,A
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: THOMPSON,JEFFREY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 12/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: CHISHOLM

DATE/TIME OF TEST: 01/17/2018 02:18:52

Diagnostics Check		OK	02:12
Air Blank		0.000	02:13
Control Test		0.080	02:13
Air Blank		0.000	02:13
Subject Sample	# 1	0.190	02:14
Breath Volume		1.191	02:14
Air Blank		0.000	02:15
Air Blank		0.000	02:16
Subject Sample	# 2	0.190	02:17
Breath Volume		1.843	02:17
Air Blank		0.000	02:17
Control Test		0.079	02:18
Air Blank		0.000	02:18
Diagnostics Check		OK	02:18

Data Download Date/Time: 01/30/2018 10:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,G,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,G
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 12/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: CARDOZA-MORENO

DATE/TIME OF TEST: 01/18/2018 21:30:11

Diagnostics Check		OK	21:23
Air Blank		0.000	21:23
Control Test		0.080	21:23
Air Blank		0.000	21:24
Subject Sample	# 1	0.147	21:25
Breath Volume		1.300	21:25
Air Blank		0.000	21:26
Air Blank		0.000	21:28
Subject Sample	# 2	0.148	21:28
Breath Volume		1.343	21:28
Air Blank		0.000	21:29
Control Test		0.079	21:29
Air Blank		0.000	21:30
Diagnostics Check		OK	21:30

Data Download Date/Time: 01/30/2018 10:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,G,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,G
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 12/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: DOW

DATE/TIME OF TEST: 01/24/2018 23:29:29

Diagnostics Check		OK	23:23
Air Blank		0.000	23:23
Control Test		0.080	23:23
Air Blank		0.000	23:24
Subject Sample	# 1	0.187	23:25
Breath Volume		1.753	23:25
Air Blank		0.000	23:25
Air Blank		0.000	23:27
Subject Sample	# 2	0.178	23:27
Breath Volume		1.679	23:27
Air Blank		0.000	23:28
Control Test		0.081	23:28
Air Blank		0.000	23:29
Diagnostics Check		OK	23:29

Data Download Date/Time: 01/30/2018 10:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TERRY,J
ARRESTING OFFICER AGENCY: Macdill Air Force Ba
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:42

SUBJECT'S LAST NAME: HOFFMAN

DATE/TIME OF TEST: 01/04/2018 19:11:10

Diagnostics Check		OK	19:02
Air Blank		0.000	19:03
Control Test		0.081	19:03
Air Blank		0.000	19:04
Subject Sample	# 1	0.178	19:06
Breath Volume		1.523	19:06
Air Blank		0.000	19:06
Air Blank		0.000	19:08
Subject Sample	# 2	0.181	19:09
Breath Volume		2.312	19:09
Air Blank		0.000	19:10
Control Test		0.078	19:10
Air Blank		0.000	19:11
Diagnostics Check		OK	19:11

Data Download Date/Time: 02/13/2018 17:45

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PECYLAK,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: MARIANI

DATE/TIME OF TEST: 01/05/2018 02:55:23

Diagnostics Check		OK	02:48
Air Blank		0.000	02:49
Control Test		0.081	02:49
Air Blank		0.000	02:50
Subject Sample	# 1	0.164	02:50
Breath Volume		1.273	02:50
Air Blank		0.000	02:51
Air Blank		0.000	02:53
Subject Sample	# 2	0.165	02:53
Breath Volume		1.683	02:53
Air Blank		0.000	02:54
Control Test		0.079	02:54
Air Blank		0.000	02:55
Diagnostics Check		OK	02:55

Data Download Date/Time: 02/13/2018 17:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:34

SUBJECT'S LAST NAME: LAMBERT

DATE/TIME OF TEST: 01/05/2018 21:01:42

Diagnostics Check		OK	20:54
Air Blank		0.000	20:55
Control Test		0.081	20:55
Air Blank		0.000	20:56
Subject Sample	# 1	0.117	20:56
Breath Volume		2.023	20:56
Air Blank		0.000	20:57
Air Blank		0.000	20:59
Subject Sample	# 2	0.118	21:00
Breath Volume		4.328	21:00
Air Blank		0.000	21:00
Control Test		0.080	21:01
Air Blank		0.000	21:01
Diagnostics Check		OK	21:01

Data Download Date/Time: 02/13/2018 17:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 01/06/2018 01:36:19

Diagnostics Check		OK	01:29
Air Blank		0.000	01:29
Control Test		0.080	01:30
Air Blank		0.000	01:30
Subject Sample	# 1	0.199	01:31
Breath Volume		1.378	01:31
Air Blank		0.000	01:31
Air Blank		0.000	01:33
Subject Sample	# 2	0.189	01:34
Breath Volume		1.425	01:34
Air Blank		0.000	01:35
Control Test		0.080	01:35
Air Blank		0.000	01:36
Diagnostics Check		OK	01:36

Data Download Date/Time: 02/13/2018 17:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HARRISON,G
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:02

SUBJECT'S LAST NAME: HEVERLING

DATE/TIME OF TEST: 01/06/2018 05:39:41

Diagnostics Check		OK	05:22	
Air Blank		0.000	05:22	
Control Test		0.080	05:23	
Air Blank		0.000	05:23	
Subject Sample	# 1	0.150	05:26	
Breath Volume		1.281	05:26	
Air Blank		0.000	05:27	
Air Blank		0.000	05:29	
Subject Sample	# 2	0.126	05:32	Volume Not Met
Breath Volume		0.144	05:32	
Air Blank		0.000	05:33	
Air Blank		0.000	05:34	
Subject Sample	# 3	0.126	05:38	Volume Not Met
Breath Volume		0.035	05:38	
Air Blank		0.000	05:38	
Control Test		0.078	05:39	
Air Blank		0.000	05:39	
Diagnostics Check		OK	05:39	

Data Download Date/Time: 02/13/2018 17:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HARRISON,G
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:02

SUBJECT'S LAST NAME: HEVERLING

DATE/TIME OF TEST: 01/06/2018 05:55:16

Diagnostics Check		OK	05:42	
Air Blank		0.000	05:42	
Control Test		0.078	05:42	
Air Blank		0.000	05:43	
Subject Sample	# 1	0.168	05:44	
Breath Volume		1.886	05:44	
Air Blank		0.000	05:45	
Air Blank		0.000	05:47	
Subject Sample	# 2	0.154	05:50	Volume Not Met
Breath Volume		0.105	05:50	
Air Blank		0.000	05:50	
Air Blank		0.000	05:52	
Subject Sample	# 3	0.161	05:53	
Breath Volume		2.183	05:53	
Air Blank		0.000	05:54	
Control Test		0.077	05:54	
Air Blank		0.000	05:55	
Diagnostics Check		OK	05:55	

Data Download Date/Time: 02/13/2018 17:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,ANGELA,V
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DAWSY,BRIAN
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 13:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:22

SUBJECT'S LAST NAME: OTTO

DATE/TIME OF TEST: 01/06/2018 14:49:25

Diagnostics Check		OK	14:43
Air Blank		0.000	14:43
Control Test		0.080	14:43
Air Blank		0.000	14:44
Subject Sample	# 1	0.217	14:44
Breath Volume		2.976	14:44
Air Blank		0.000	14:45
Air Blank		0.000	14:47
Subject Sample	# 2	0.226	14:47
Breath Volume		3.750	14:47
Air Blank		0.000	14:48
Control Test		0.079	14:48
Air Blank		0.000	14:49
Diagnostics Check		OK	14:49

Data Download Date/Time: 02/13/2018 17:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS,RONALD,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DEJESUS,V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: MARTINEZ CONTRERAS

DATE/TIME OF TEST: 01/07/2018 00:11:47

Diagnostics Check		OK	00:05
Air Blank		0.000	00:06
Control Test		0.080	00:06
Air Blank		0.000	00:06
Subject Sample	# 1	0.107	00:07
Breath Volume		1.839	00:07
Air Blank		0.000	00:07
Air Blank		0.000	00:09
Subject Sample	# 2	0.105	00:10
Breath Volume		1.527	00:10
Air Blank		0.000	00:10
Control Test		0.080	00:11
Air Blank		0.000	00:11
Diagnostics Check		OK	00:11

Data Download Date/Time: 02/13/2018 17:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: HENRIQUEZ DIAZ

DATE/TIME OF TEST: 01/07/2018 03:01:26

Diagnostics Check		OK	02:52
Air Blank		0.000	02:53
Control Test		0.080	02:53
Air Blank		0.000	02:53
Subject Sample	# 1	0.182	02:56
Breath Volume		1.421	02:56
Air Blank		0.000	02:57
Air Blank		0.000	02:59
Subject Sample	# 2	0.183	02:59
Breath Volume		1.980	02:59
Air Blank		0.000	03:00
Control Test		0.078	03:00
Air Blank		0.000	03:01
Diagnostics Check		OK	03:01

Data Download Date/Time: 02/13/2018 17:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HONEYGHON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: WARD

DATE/TIME OF TEST: 01/07/2018 04:54:45

Diagnostics Check		OK	04:47
Air Blank		0.000	04:48
Control Test		0.080	04:48
Air Blank		0.000	04:49
Subject Sample	# 1	0.081	04:49
Breath Volume		4.656	04:49
Air Blank		0.000	04:50
Air Blank		0.000	04:52
Subject Sample	# 2	0.084	04:53
Breath Volume		3.781	04:53
Air Blank		0.000	04:53
Control Test		0.080	04:54
Air Blank		0.000	04:54
Diagnostics Check		OK	04:54

Data Download Date/Time: 02/13/2018 17:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS,RONALD,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: CROGNALE

DATE/TIME OF TEST: 01/07/2018 22:54:31

Diagnostics Check		OK	22:45
Air Blank		0.000	22:46
Control Test		0.080	22:46
Air Blank		0.000	22:47
Subject Sample	# 1	0.061	22:47
Breath Volume		1.500	22:47
Air Blank		0.000	22:48
Air Blank		0.000	22:50
Subject Sample	# 2	0.061	22:52
Breath Volume		1.257	22:52
Air Blank		0.000	22:53
Control Test		0.081	22:53
Air Blank		0.000	22:54
Diagnostics Check		OK	22:54

Data Download Date/Time: 02/13/2018 17:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SINANI,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 01/08/2018 05:02:00

Diagnostics Check		OK	04:55
Air Blank		0.000	04:56
Control Test		0.080	04:56
Air Blank		0.000	04:57
Subject Sample	# 1	0.158	04:57
Breath Volume		1.863	04:57
Air Blank		0.000	04:58
Air Blank		0.000	05:00
Subject Sample	# 2	0.165	05:00
Breath Volume		1.199	05:00
Air Blank		0.000	05:01
Control Test		0.080	05:01
Air Blank		0.000	05:01
Diagnostics Check		OK	05:02

Data Download Date/Time: 02/13/2018 17:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS,RONALD,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: JUDD

DATE/TIME OF TEST: 01/08/2018 21:33:18

Diagnostics Check		OK	21:27
Air Blank		0.000	21:27
Control Test		0.079	21:27
Air Blank		0.000	21:28
Subject Sample	# 1	0.133	21:28
Breath Volume		1.343	21:28
Air Blank		0.000	21:29
Air Blank		0.000	21:31
Subject Sample	# 2	0.135	21:31
Breath Volume		1.734	21:31
Air Blank		0.000	21:32
Control Test		0.079	21:32
Air Blank		0.000	21:33
Diagnostics Check		OK	21:33

Data Download Date/Time: 02/13/2018 17:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: BOYD

DATE/TIME OF TEST: 01/09/2018 04:15:26

Diagnostics Check		OK	04:09
Air Blank		0.000	04:09
Control Test		0.079	04:09
Air Blank		0.000	04:10
Subject Sample	# 1	0.216	04:10
Breath Volume		3.281	04:10
Air Blank		0.000	04:11
Air Blank		0.000	04:13
Subject Sample	# 2	0.226	04:13
Breath Volume		3.316	04:13
Air Blank		0.000	04:14
Control Test		0.077	04:14
Air Blank		0.000	04:15
Diagnostics Check		OK	04:15

Data Download Date/Time: 02/13/2018 17:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: COSTA,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: HUDSON

DATE/TIME OF TEST: 01/10/2018 04:21:00

Diagnostics Check		OK	04:13
Air Blank		0.000	04:14
Control Test		0.080	04:14
Air Blank		0.000	04:15
Subject Sample	# 1	0.000	04:15
Breath Volume		1.242	04:15
Air Blank		0.000	04:16
Air Blank		0.000	04:18
Subject Sample	# 2	0.000	04:19
Breath Volume		1.378	04:19
Air Blank		0.000	04:20
Control Test		0.080	04:20
Air Blank		0.000	04:20
Diagnostics Check		OK	04:21

Data Download Date/Time: 02/13/2018 17:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS,RONALD,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: BORRELI

DATE/TIME OF TEST: 01/12/2018 00:35:39

Diagnostics Check		OK	00:26	
Air Blank		0.000	00:27	
Control Test		0.081	00:27	
Air Blank		0.000	00:28	
Subject Sample	# 1		00:31	No Sample Provided
Breath Volume		0.000	00:31	
Air Blank		0.000	00:31	
Air Blank		0.000	00:33	
Subject Sample	# 2		00:34	Interferent Detect
Breath Volume		1.945	00:34	
Air Blank		0.000	00:34	
Control Test		0.080	00:35	
Air Blank		0.000	00:35	
Diagnostics Check		OK	00:35	

Data Download Date/Time: 02/13/2018 17:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS,RONALD,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: BORRELI

DATE/TIME OF TEST: 01/12/2018 00:41:37

Diagnostics Check		OK	00:38	
Air Blank		0.000	00:38	
Control Test		0.079	00:38	
Air Blank		0.000	00:39	
Subject Sample	# 1		00:40	Interferent Detect
Breath Volume		1.406	00:40	
Air Blank		0.000	00:40	
Control Test		0.079	00:41	
Air Blank		0.000	00:41	
Diagnostics Check		OK	00:41	

Data Download Date/Time: 02/13/2018 17:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS,RONALD,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRIGHTWELL,R
ARRESTING OFFICER AGENCY: U of So FL Campus PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: PRINZ

DATE/TIME OF TEST: 01/12/2018 02:46:59

Diagnostics Check		OK	02:40
Air Blank		0.000	02:41
Control Test		0.079	02:41
Air Blank		0.000	02:41
Subject Sample	# 1	0.000	02:42
Breath Volume		3.308	02:42
Air Blank		0.000	02:43
Air Blank		0.000	02:45
Subject Sample	# 2	0.000	02:45
Breath Volume		4.035	02:45
Air Blank		0.000	02:46
Control Test		0.080	02:46
Air Blank		0.000	02:46
Diagnostics Check		OK	02:46

Data Download Date/Time: 02/13/2018 17:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SHEINBERG,T
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:01

SUBJECT'S LAST NAME: FERMIN

DATE/TIME OF TEST: 01/12/2018 06:29:06

Diagnostics Check		OK	06:21
Air Blank		0.000	06:22
Control Test		0.079	06:22
Air Blank		0.000	06:22
Subject Sample	# 1	0.159	06:24
Breath Volume		2.777	06:24
Air Blank		0.000	06:24
Air Blank		0.000	06:26
Subject Sample	# 2	0.164	06:27
Breath Volume		2.628	06:27
Air Blank		0.000	06:28
Control Test		0.080	06:28
Air Blank		0.000	06:29
Diagnostics Check		OK	06:29

Data Download Date/Time: 02/13/2018 17:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: VELEZ

DATE/TIME OF TEST: 01/12/2018 21:50:26

Diagnostics Check		OK	21:42
Air Blank		0.000	21:43
Control Test		0.080	21:43
Air Blank		0.000	21:44
Subject Sample	# 1	0.174	21:45
Breath Volume		2.699	21:45
Air Blank		0.000	21:45
Air Blank		0.000	21:47
Subject Sample	# 2	0.163	21:48
Breath Volume		1.566	21:48
Air Blank		0.000	21:49
Control Test		0.078	21:49
Air Blank		0.000	21:50
Diagnostics Check		OK	21:50

Data Download Date/Time: 02/13/2018 17:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS,RONALD,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: ASENCIO

DATE/TIME OF TEST: 01/13/2018 01:44:02

Diagnostics Check		OK	01:38	
Air Blank		0.000	01:38	
Control Test		0.080	01:39	
Air Blank		0.000	01:39	
Subject Sample	# 1		01:42	Subject Test Refused
Breath Volume		0.000	01:42	
Air Blank		0.000	01:43	
Control Test		0.079	01:43	
Air Blank		0.000	01:43	
Diagnostics Check		OK	01:44	

Data Download Date/Time: 02/13/2018 17:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CONVERSE,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:28

SUBJECT'S LAST NAME: CANO

DATE/TIME OF TEST: 01/13/2018 05:54:14

Diagnostics Check		OK	05:48
Air Blank		0.000	05:48
Control Test		0.079	05:48
Air Blank		0.000	05:49
Subject Sample	# 1	0.117	05:49
Breath Volume		3.207	05:49
Air Blank		0.000	05:50
Air Blank		0.000	05:52
Subject Sample	# 2	0.117	05:52
Breath Volume		3.292	05:52
Air Blank		0.000	05:53
Control Test		0.079	05:53
Air Blank		0.000	05:54
Diagnostics Check		OK	05:54

Data Download Date/Time: 02/13/2018 17:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WOODEN,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: CARTWRIGHT

DATE/TIME OF TEST: 01/13/2018 23:06:43

Diagnostics Check		OK	23:00
Air Blank		0.000	23:00
Control Test		0.080	23:00
Air Blank		0.000	23:01
Subject Sample	# 1	0.140	23:02
Breath Volume		2.148	23:02
Air Blank		0.000	23:02
Air Blank		0.000	23:04
Subject Sample	# 2	0.145	23:05
Breath Volume		3.523	23:05
Air Blank		0.000	23:05
Control Test		0.079	23:06
Air Blank		0.000	23:06
Diagnostics Check		OK	23:06

Data Download Date/Time: 02/13/2018 17:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: COSTA,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: HIRD

DATE/TIME OF TEST: 01/14/2018 01:53:56

Diagnostics Check		OK	01:46
Air Blank		0.000	01:47
Control Test		0.080	01:47
Air Blank		0.000	01:48
Subject Sample	# 1	0.159	01:48
Breath Volume		1.539	01:48
Air Blank		0.000	01:49
Air Blank		0.000	01:51
Subject Sample	# 2	0.152	01:52
Breath Volume		1.382	01:52
Air Blank		0.000	01:53
Control Test		0.080	01:53
Air Blank		0.000	01:53
Diagnostics Check		OK	01:53

Data Download Date/Time: 02/13/2018 17:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: BEARDSLEY

DATE/TIME OF TEST: 01/14/2018 03:33:56

Diagnostics Check		OK	03:27
Air Blank		0.000	03:28
Control Test		0.080	03:28
Air Blank		0.000	03:28
Subject Sample	# 1	0.171	03:29
Breath Volume		2.390	03:29
Air Blank		0.000	03:30
Air Blank		0.000	03:31
Subject Sample	# 2	0.161	03:32
Breath Volume		2.039	03:32
Air Blank		0.000	03:33
Control Test		0.080	03:33
Air Blank		0.000	03:33
Diagnostics Check		OK	03:33

Data Download Date/Time: 02/13/2018 17:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,REBECCA,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:24

SUBJECT'S LAST NAME: MEISEL

DATE/TIME OF TEST: 01/14/2018 18:54:51

Diagnostics Check		OK	18:44
Air Blank		0.000	18:45
Control Test		0.080	18:45
Air Blank		0.000	18:46
Subject Sample	# 1	0.498	18:47
Breath Volume		1.449	18:47
Air Blank		0.000	18:47
Air Blank		0.000	18:49
Subject Sample	# 2	0.475	18:49
Breath Volume		1.507	18:49
Air Blank		0.000	18:50
Air Blank		0.000	18:52
Subject Sample	# 3	0.470	18:53
Breath Volume		2.347	18:53
Air Blank		0.000	18:53
Control Test		0.078	18:54
Air Blank		0.000	18:54
Diagnostics Check		OK	18:54

Data Download Date/Time: 02/13/2018 17:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: WATTS

DATE/TIME OF TEST: 01/15/2018 04:45:36

Diagnostics Check		OK	04:37
Air Blank		0.000	04:38
Control Test		0.080	04:38
Air Blank		0.000	04:38
Subject Sample	# 1	0.173	04:39
Breath Volume		1.964	04:39
Air Blank		0.000	04:40
Air Blank		0.000	04:41
Subject Sample	# 2	0.160	04:43
Breath Volume		1.328	04:43
Air Blank		0.000	04:44
Control Test		0.081	04:44
Air Blank		0.000	04:45
Diagnostics Check		OK	04:45

Data Download Date/Time: 02/13/2018 17:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS,RONALD,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRONSON,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: REYNOLDS

DATE/TIME OF TEST: 01/16/2018 22:03:38

Diagnostics Check		OK	21:56
Air Blank		0.000	21:57
Control Test		0.080	21:57
Air Blank		0.000	21:58
Subject Sample	# 1	0.203	21:59
Breath Volume		1.347	21:59
Air Blank		0.000	21:59
Air Blank		0.000	22:01
Subject Sample	# 2	0.214	22:02
Breath Volume		1.304	22:02
Air Blank		0.000	22:02
Control Test		0.078	22:03
Air Blank		0.000	22:03
Diagnostics Check		OK	22:03

Data Download Date/Time: 02/13/2018 17:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS,RONALD,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARRELL,Q
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: VELEZ

DATE/TIME OF TEST: 01/17/2018 23:03:02

Diagnostics Check		OK	22:56
Air Blank		0.000	22:57
Control Test		0.080	22:57
Air Blank		0.000	22:57
Subject Sample	# 1	0.187	22:58
Breath Volume		1.917	22:58
Air Blank		0.000	22:59
Air Blank		0.000	23:01
Subject Sample	# 2	0.195	23:01
Breath Volume		1.769	23:01
Air Blank		0.000	23:02
Control Test		0.079	23:02
Air Blank		0.000	23:02
Diagnostics Check		OK	23:03

Data Download Date/Time: 02/13/2018 17:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:47

SUBJECT'S LAST NAME: MASSINGILL

DATE/TIME OF TEST: 01/18/2018 19:14:29

Diagnostics Check		OK	19:07
Air Blank		0.000	19:08
Control Test		0.081	19:08
Air Blank		0.000	19:09
Subject Sample	# 1	0.222	19:09
Breath Volume		1.429	19:09
Air Blank		0.000	19:10
Air Blank		0.000	19:12
Subject Sample	# 2	0.220	19:12
Breath Volume		1.613	19:12
Air Blank		0.000	19:13
Control Test		0.081	19:13
Air Blank		0.000	19:14
Diagnostics Check		OK	19:14

Data Download Date/Time: 02/13/2018 17:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MACIAS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: LEAL-RAMIREZ

DATE/TIME OF TEST: 01/19/2018 23:50:02

Diagnostics Check		OK	23:42
Air Blank		0.000	23:42
Control Test		0.081	23:42
Air Blank		0.000	23:43
Subject Sample	# 1	0.097	23:45
Breath Volume		1.222	23:45
Air Blank		0.000	23:45
Air Blank		0.000	23:47
Subject Sample	# 2	0.106	23:48
Breath Volume		1.566	23:48
Air Blank		0.000	23:49
Control Test		0.080	23:49
Air Blank		0.000	23:49
Diagnostics Check		OK	23:50

Data Download Date/Time: 02/13/2018 17:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LYON,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: KOLCZYNSKI

DATE/TIME OF TEST: 01/20/2018 01:09:28

Diagnostics Check		OK	01:06	
Air Blank		0.000	01:06	
Control Test		0.081	01:06	
Air Blank		0.000	01:07	
Subject Sample	# 1		01:08	Subject Test Refused
Breath Volume		0.000	01:08	
Air Blank		0.000	01:08	
Control Test		0.080	01:08	
Air Blank		0.000	01:09	
Diagnostics Check		OK	01:09	

Data Download Date/Time: 02/13/2018 17:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: MARSHALL

DATE/TIME OF TEST: 01/20/2018 04:37:51

Diagnostics Check		OK	04:31
Air Blank		0.000	04:31
Control Test		0.080	04:31
Air Blank		0.000	04:32
Subject Sample	# 1	0.252	04:32
Breath Volume		1.269	04:32
Air Blank		0.000	04:33
Air Blank		0.000	04:35
Subject Sample	# 2	0.260	04:36
Breath Volume		2.011	04:36
Air Blank		0.000	04:36
Control Test		0.080	04:37
Air Blank		0.000	04:37
Diagnostics Check		OK	04:37

Data Download Date/Time: 02/13/2018 17:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,ANGELA,V
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CAMPBELL,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 12:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:28

SUBJECT'S LAST NAME: ARNETT

DATE/TIME OF TEST: 01/20/2018 13:55:22

Diagnostics Check		OK	13:49
Air Blank		0.000	13:49
Control Test		0.080	13:50
Air Blank		0.000	13:50
Subject Sample	# 1	0.000	13:51
Breath Volume		2.382	13:51
Air Blank		0.000	13:51
Air Blank		0.000	13:53
Subject Sample	# 2	0.000	13:54
Breath Volume		2.988	13:54
Air Blank		0.000	13:54
Control Test		0.080	13:54
Air Blank		0.000	13:55
Diagnostics Check		OK	13:55

Data Download Date/Time: 02/13/2018 17:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: BAUTISTA MAYORQUIN

DATE/TIME OF TEST: 01/20/2018 23:33:43

Diagnostics Check		OK	23:25
Air Blank		0.000	23:25
Control Test		0.080	23:26
Air Blank		0.000	23:26
Subject Sample	# 1	0.232	23:29
Breath Volume		2.210	23:29
Air Blank		0.000	23:29
Air Blank		0.000	23:31
Subject Sample	# 2	0.250	23:32
Breath Volume		1.335	23:32
Air Blank		0.000	23:32
Control Test		0.078	23:33
Air Blank		0.000	23:33
Diagnostics Check		OK	23:33

Data Download Date/Time: 02/13/2018 17:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: LELAULU

DATE/TIME OF TEST: 01/21/2018 02:04:14

Diagnostics Check		OK	01:56
Air Blank		0.000	01:57
Control Test		0.080	01:57
Air Blank		0.000	01:57
Subject Sample	# 1	0.221	01:58
Breath Volume		1.535	01:58
Air Blank		0.000	01:59
Air Blank		0.000	02:01
Subject Sample	# 2	0.235	02:02
Breath Volume		1.710	02:02
Air Blank		0.000	02:03
Control Test		0.078	02:03
Air Blank		0.000	02:04
Diagnostics Check		OK	02:04

Data Download Date/Time: 02/13/2018 17:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NIEUWENDAM,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: FUENTES

DATE/TIME OF TEST: 01/21/2018 02:29:41

Diagnostics Check		OK	02:23
Air Blank		0.000	02:23
Control Test		0.079	02:24
Air Blank		0.000	02:24
Subject Sample	# 1	0.000	02:25
Breath Volume		3.671	02:25
Air Blank		0.000	02:25
Air Blank		0.000	02:27
Subject Sample	# 2	0.000	02:28
Breath Volume		3.843	02:28
Air Blank		0.000	02:28
Control Test		0.079	02:29
Air Blank		0.000	02:29
Diagnostics Check		OK	02:29

Data Download Date/Time: 02/13/2018 17:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NIEUWENDAM,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:34

SUBJECT'S LAST NAME: MYRICK II

DATE/TIME OF TEST: 01/21/2018 06:02:58

Diagnostics Check		OK	05:54
Air Blank		0.000	05:54
Control Test		0.080	05:54
Air Blank		0.000	05:55
Subject Sample	# 1	0.109	05:57
Breath Volume		2.097	05:57
Air Blank		0.000	05:58
Air Blank		0.000	06:00
Subject Sample	# 2	0.111	06:01
Breath Volume		2.113	06:01
Air Blank		0.000	06:02
Control Test		0.079	06:02
Air Blank		0.000	06:02
Diagnostics Check		OK	06:02

Data Download Date/Time: 02/13/2018 17:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS,RONALD,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DEJESUS,V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:07

SUBJECT'S LAST NAME: GARCIA-CASTRO

DATE/TIME OF TEST: 01/21/2018 21:34:40

Diagnostics Check		OK	21:27
Air Blank		0.000	21:28
Control Test		0.080	21:28
Air Blank		0.000	21:29
Subject Sample	# 1	0.226	21:30
Breath Volume		1.453	21:30
Air Blank		0.000	21:30
Air Blank		0.000	21:32
Subject Sample	# 2	0.219	21:33
Breath Volume		1.214	21:33
Air Blank		0.000	21:33
Control Test		0.078	21:34
Air Blank		0.000	21:34
Diagnostics Check		OK	21:34

Data Download Date/Time: 02/13/2018 17:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: WOMACK

DATE/TIME OF TEST: 01/22/2018 03:58:11

Diagnostics Check		OK	03:46	
Air Blank		0.000	03:47	
Control Test		0.079	03:47	
Air Blank		0.000	03:48	
Subject Sample	# 1		03:51	No Sample Provided
Breath Volume		0.000	03:51	
Air Blank		0.000	03:51	
Air Blank		0.000	03:53	
Subject Sample	# 2		03:56	No Sample Provided
Breath Volume		0.000	03:56	
Air Blank		0.000	03:57	
Control Test		0.079	03:57	
Air Blank		0.000	03:58	
Diagnostics Check		OK	03:58	

Data Download Date/Time: 02/13/2018 17:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKIPPER,ROGER,G
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TEST,TEST
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: SYS CHECK
ARREST TIME: 08:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:00

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 01/22/2018 09:24:28

Diagnostics Check		OK	09:22	
Air Blank		0.000	09:22	
Control Test		0.079	09:23	
Air Blank		0.000	09:23	
Subject Sample	# 1		09:24	Sequence Aborted
Breath Volume		0.000	09:24	
Air Blank		0.000	09:24	

Data Download Date/Time: 02/13/2018 17:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PECYLAK,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: CRANFORD III

DATE/TIME OF TEST: 01/23/2018 00:03:58

Diagnostics Check		OK	23:57
Air Blank		0.000	23:57
Control Test		0.079	23:58
Air Blank		0.000	23:58
Subject Sample	# 2	0.200	23:59
Breath Volume		1.937	23:59
Air Blank		0.000	23:59
Air Blank		0.000	00:01
Subject Sample	# 1	0.200	00:02
Breath Volume		1.792	00:02
Air Blank		0.000	00:03
Control Test		0.079	00:03
Air Blank		0.000	00:03
Diagnostics Check		OK	00:03

Data Download Date/Time: 02/13/2018 17:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKIPPER,ROGER,G
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TEST,TEST
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: SYS CHECK
ARREST TIME: 11:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:00

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 01/23/2018 13:46:03

Diagnostics Check		OK	13:40
Air Blank		0.000	13:40
Control Test		0.080	13:40
Air Blank		0.000	13:41
Subject Sample	# 1	0.000	13:41
Breath Volume		1.539	13:41
Air Blank		0.000	13:42
Air Blank		0.000	13:44
Subject Sample	# 2	0.000	13:44
Breath Volume		1.386	13:44
Air Blank		0.000	13:45
Control Test		0.080	13:45
Air Blank		0.000	13:45
Diagnostics Check		OK	13:46

Data Download Date/Time: 02/13/2018 17:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIMPKINS,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11117080A3
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: GONTARSKI

DATE/TIME OF TEST: 01/04/2018 23:44:37

Diagnostics Check		OK	23:38
Air Blank		0.000	23:38
Control Test		0.083	23:39
Air Blank		0.000	23:39
Subject Sample	# 1	0.110	23:40
Breath Volume		3.121	23:40
Air Blank		0.000	23:40
Air Blank		0.000	23:42
Subject Sample	# 2	0.118	23:43
Breath Volume		2.355	23:43
Air Blank		0.000	23:43
Control Test		0.081	23:44
Air Blank		0.000	23:44
Diagnostics Check		OK	23:44

Data Download Date/Time: 02/13/2018 17:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11117080A3
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: ARANDA TRUJILLO

DATE/TIME OF TEST: 01/05/2018 02:54:54

Diagnostics Check		OK	02:48
Air Blank		0.000	02:48
Control Test		0.082	02:49
Air Blank		0.000	02:49
Subject Sample	# 1	0.133	02:50
Breath Volume		2.105	02:50
Air Blank		0.000	02:50
Air Blank		0.000	02:52
Subject Sample	# 2	0.136	02:53
Breath Volume		2.390	02:53
Air Blank		0.000	02:53
Control Test		0.082	02:54
Air Blank		0.000	02:54
Diagnostics Check		OK	02:54

Data Download Date/Time: 02/13/2018 17:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NIEUWENDAM,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11117080A3
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: FLYNN

DATE/TIME OF TEST: 01/05/2018 22:56:59

Diagnostics Check		OK	22:50
Air Blank		0.000	22:50
Control Test		0.081	22:51
Air Blank		0.000	22:51
Subject Sample	# 1	0.000	22:52
Breath Volume		2.550	22:52
Air Blank		0.000	22:53
Air Blank		0.000	22:55
Subject Sample	# 2	0.000	22:55
Breath Volume		2.144	22:55
Air Blank		0.000	22:56
Control Test		0.082	22:56
Air Blank		0.000	22:56
Diagnostics Check		OK	22:56

Data Download Date/Time: 02/13/2018 17:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11117080A3
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 01/06/2018 02:09:56

Diagnostics Check		OK	02:03
Air Blank		0.000	02:03
Control Test		0.082	02:04
Air Blank		0.000	02:04
Subject Sample	# 1	0.150	02:05
Breath Volume		4.101	02:05
Air Blank		0.000	02:05
Air Blank		0.000	02:07
Subject Sample	# 2	0.145	02:08
Breath Volume		3.574	02:08
Air Blank		0.000	02:09
Control Test		0.080	02:09
Air Blank		0.000	02:09
Diagnostics Check		OK	02:09

Data Download Date/Time: 02/13/2018 17:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SULLIVAN,T
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11117080A3
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:06

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 01/06/2018 05:32:28

Diagnostics Check		OK	05:26
Air Blank		0.000	05:26
Control Test		0.081	05:26
Air Blank		0.000	05:27
Subject Sample	# 1	0.099	05:27
Breath Volume		2.847	05:27
Air Blank		0.000	05:28
Air Blank		0.000	05:30
Subject Sample	# 2	0.097	05:30
Breath Volume		3.199	05:30
Air Blank		0.000	05:31
Control Test		0.081	05:31
Air Blank		0.000	05:32
Diagnostics Check		OK	05:32

Data Download Date/Time: 02/13/2018 17:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SHEINBERG,T
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11117080A3
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: JULIO DOMINGO

DATE/TIME OF TEST: 01/06/2018 23:26:41

Diagnostics Check		OK	23:20
Air Blank		0.000	23:20
Control Test		0.082	23:20
Air Blank		0.000	23:21
Subject Sample	# 1	0.053	23:22
Breath Volume		1.257	23:22
Air Blank		0.000	23:22
Air Blank		0.000	23:24
Subject Sample	# 2	0.051	23:25
Breath Volume		1.269	23:25
Air Blank		0.000	23:25
Control Test		0.082	23:26
Air Blank		0.000	23:26
Diagnostics Check		OK	23:26

Data Download Date/Time: 02/13/2018 17:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS,RONALD,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11117080A3
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: BURGOS

DATE/TIME OF TEST: 01/07/2018 03:53:18

Diagnostics Check		OK	03:44
Air Blank		0.000	03:45
Control Test		0.081	03:45
Air Blank		0.000	03:45
Subject Sample	# 1	0.208	03:48
Breath Volume		2.152	03:48
Air Blank		0.000	03:49
Air Blank		0.000	03:51
Subject Sample	# 2	0.215	03:51
Breath Volume		2.476	03:51
Air Blank		0.000	03:52
Control Test		0.078	03:52
Air Blank		0.000	03:53
Diagnostics Check		OK	03:53

Data Download Date/Time: 02/13/2018 17:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS,RONALD,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11117080A3
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:08

SUBJECT'S LAST NAME: GUIDOTTI

DATE/TIME OF TEST: 01/07/2018 05:35:10

Diagnostics Check		OK	05:28
Air Blank		0.000	05:29
Control Test		0.082	05:29
Air Blank		0.000	05:29
Subject Sample	# 1	0.000	05:30
Breath Volume		1.480	05:30
Air Blank		0.000	05:31
Air Blank		0.000	05:33
Subject Sample	# 2	0.000	05:33
Breath Volume		1.667	05:33
Air Blank		0.000	05:34
Control Test		0.081	05:34
Air Blank		0.000	05:35
Diagnostics Check		OK	05:35

Data Download Date/Time: 02/13/2018 17:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRANKLIN,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11117080A3
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:44

SUBJECT'S LAST NAME: SCHULTZ

DATE/TIME OF TEST: 01/07/2018 06:07:00

Diagnostics Check		OK	06:04	
Air Blank		0.000	06:04	
Control Test		0.081	06:04	
Air Blank		0.000	06:05	
Subject Sample	# 1		06:05	Subject Test Refused
Breath Volume		0.000	06:05	
Air Blank		0.000	06:06	
Control Test		0.081	06:06	
Air Blank		0.000	06:06	
Diagnostics Check		OK	06:07	

Data Download Date/Time: 02/13/2018 17:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11117080A3
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: GOMEZ GOMEZ

DATE/TIME OF TEST: 01/07/2018 23:19:25

Diagnostics Check		OK	23:12
Air Blank		0.000	23:13
Control Test		0.081	23:13
Air Blank		0.000	23:14
Subject Sample	# 1	0.192	23:14
Breath Volume		2.851	23:14
Air Blank		0.000	23:15
Air Blank		0.000	23:17
Subject Sample	# 2	0.188	23:17
Breath Volume		3.230	23:17
Air Blank		0.000	23:18
Control Test		0.080	23:18
Air Blank		0.000	23:19
Diagnostics Check		OK	23:19

Data Download Date/Time: 02/13/2018 17:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKIPPER,ROGER,G
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: REICHERT,C 4763
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11117080A3
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 12:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:45

SUBJECT'S LAST NAME: CREW

DATE/TIME OF TEST: 01/08/2018 13:12:04

Diagnostics Check		OK	13:05
Air Blank		0.000	13:06
Control Test		0.080	13:06
Air Blank		0.000	13:07
Subject Sample	# 1	0.016	13:07
Breath Volume		2.339	13:07
Air Blank		0.000	13:08
Air Blank		0.000	13:10
Subject Sample	# 2	0.016	13:10
Breath Volume		2.109	13:10
Air Blank		0.000	13:11
Control Test		0.081	13:11
Air Blank		0.000	13:11
Diagnostics Check		OK	13:12

Data Download Date/Time: 02/13/2018 17:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PECYLAK,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11117080A3
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:02

SUBJECT'S LAST NAME: BARKER

DATE/TIME OF TEST: 01/09/2018 20:28:51

Diagnostics Check		OK	20:22
Air Blank		0.000	20:23
Control Test		0.080	20:23
Air Blank		0.000	20:23
Subject Sample	# 1	0.097	20:24
Breath Volume		2.292	20:24
Air Blank		0.000	20:24
Air Blank		0.000	20:26
Subject Sample	# 2	0.088	20:27
Breath Volume		1.535	20:27
Air Blank		0.000	20:27
Control Test		0.080	20:28
Air Blank		0.000	20:28
Diagnostics Check		OK	20:28

Data Download Date/Time: 02/13/2018 17:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HONEYGHON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11117080A3
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 01/10/2018 23:36:38

Diagnostics Check		OK	23:24	
Air Blank		0.000	23:25	
Control Test		0.080	23:25	
Air Blank		0.000	23:26	
Subject Sample	# 1	0.160	23:29	Volume Not Met
Breath Volume		0.128	23:29	
Air Blank		0.000	23:30	
Air Blank		0.000	23:31	
Subject Sample	# 2	0.159	23:34	Volume Not Met
Breath Volume		0.242	23:34	
Air Blank		0.000	23:35	
Control Test		0.078	23:35	
Air Blank		0.000	23:36	
Diagnostics Check		OK	23:36	

Data Download Date/Time: 02/13/2018 17:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HONEYGHON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11117080A3
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 01/10/2018 23:49:17

Diagnostics Check		OK	23:38
Air Blank		0.000	23:39
Control Test		0.079	23:39
Air Blank		0.000	23:40
Subject Sample	# 1	0.179	23:41
Breath Volume		1.339	23:41
Air Blank		0.000	23:42
Air Blank		0.000	23:44
Subject Sample	# 2	0.167	23:47
Breath Volume		1.160	23:47
Air Blank		0.000	23:48
Control Test		0.078	23:48
Air Blank		0.000	23:49
Diagnostics Check		OK	23:49

Data Download Date/Time: 02/13/2018 17:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,ANGELA,V
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HILES,STEPHEN
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11117080A3
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 07:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:40

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 01/11/2018 08:07:42

Diagnostics Check		OK	08:01
Air Blank		0.000	08:01
Control Test		0.080	08:01
Air Blank		0.000	08:02
Subject Sample	# 1	0.138	08:03
Breath Volume		1.960	08:03
Air Blank		0.000	08:03
Air Blank		0.000	08:05
Subject Sample	# 2	0.136	08:06
Breath Volume		2.343	08:06
Air Blank		0.000	08:06
Control Test		0.080	08:07
Air Blank		0.000	08:07
Diagnostics Check		OK	08:07

Data Download Date/Time: 02/13/2018 17:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRONSON,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11117080A3
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: BUY

DATE/TIME OF TEST: 01/12/2018 00:46:01

Diagnostics Check		OK	00:38
Air Blank		0.000	00:39
Control Test		0.080	00:39
Air Blank		0.000	00:40
Subject Sample	# 1	0.131	00:40
Breath Volume		1.105	00:40
Air Blank		0.000	00:41
Air Blank		0.000	00:43
Subject Sample	# 2	0.136	00:44
Breath Volume		1.144	00:44
Air Blank		0.000	00:45
Control Test		0.079	00:45
Air Blank		0.000	00:45
Diagnostics Check		OK	00:46

Data Download Date/Time: 02/13/2018 17:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS,RONALD,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SULLIVAN,T
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11117080A3
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: SIFF

DATE/TIME OF TEST: 01/12/2018 05:16:28

Diagnostics Check		OK	05:05
Air Blank		0.000	05:06
Control Test		0.081	05:06
Air Blank		0.000	05:06
Subject Sample	# 1	0.172	05:09
Breath Volume		1.140	05:09
Air Blank		0.000	05:10
Air Blank		0.000	05:12
Subject Sample	# 2	0.171	05:14
Breath Volume		2.535	05:14
Air Blank		0.000	05:15
Control Test		0.079	05:15
Air Blank		0.000	05:16
Diagnostics Check		OK	05:16

Data Download Date/Time: 02/13/2018 17:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PECYLAK,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11117080A3
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: CARRENO

DATE/TIME OF TEST: 01/14/2018 00:21:45

Diagnostics Check		OK	00:14
Air Blank		0.000	00:15
Control Test		0.080	00:15
Air Blank		0.000	00:15
Subject Sample	# 1	0.167	00:17
Breath Volume		1.683	00:17
Air Blank		0.000	00:17
Air Blank		0.000	00:19
Subject Sample	# 2	0.174	00:20
Breath Volume		1.300	00:20
Air Blank		0.000	00:20
Control Test		0.080	00:21
Air Blank		0.000	00:21
Diagnostics Check		OK	00:21

Data Download Date/Time: 02/13/2018 17:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIMS,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11117080A3
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: SNYDER

DATE/TIME OF TEST: 01/14/2018 03:47:01

Diagnostics Check		OK	03:40
Air Blank		0.000	03:41
Control Test		0.081	03:41
Air Blank		0.000	03:41
Subject Sample	# 1	0.182	03:42
Breath Volume		2.960	03:42
Air Blank		0.000	03:43
Air Blank		0.000	03:44
Subject Sample	# 2	0.171	03:45
Breath Volume		1.597	03:45
Air Blank		0.000	03:46
Control Test		0.081	03:46
Air Blank		0.000	03:46
Diagnostics Check		OK	03:47

Data Download Date/Time: 02/13/2018 17:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NIEUWENDAM,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11117080A3
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: CARLISI IV

DATE/TIME OF TEST: 01/14/2018 04:40:37

Diagnostics Check		OK	04:31
Air Blank		0.000	04:31
Control Test		0.081	04:32
Air Blank		0.000	04:32
Subject Sample	# 1	0.142	04:33
Breath Volume		1.113	04:33
Air Blank		0.000	04:34
Air Blank		0.000	04:36
Subject Sample	# 2	0.144	04:39
Breath Volume		1.703	04:39
Air Blank		0.000	04:39
Control Test		0.080	04:40
Air Blank		0.000	04:40
Diagnostics Check		OK	04:40

Data Download Date/Time: 02/13/2018 17:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11117080A3
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: MCDONALD

DATE/TIME OF TEST: 01/14/2018 19:58:09

Diagnostics Check		OK	19:50
Air Blank		0.000	19:51
Control Test		0.081	19:51
Air Blank		0.000	19:51
Subject Sample	# 1	0.238	19:52
Breath Volume		1.253	19:52
Air Blank		0.000	19:53
Air Blank		0.000	19:55
Subject Sample	# 2	0.238	19:56
Breath Volume		1.550	19:56
Air Blank		0.000	19:57
Control Test		0.080	19:57
Air Blank		0.000	19:58
Diagnostics Check		OK	19:58

Data Download Date/Time: 02/13/2018 17:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11117080A3
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: MEEK

DATE/TIME OF TEST: 01/15/2018 04:13:32

Diagnostics Check		OK	04:07
Air Blank		0.000	04:07
Control Test		0.082	04:07
Air Blank		0.000	04:08
Subject Sample	# 1	0.221	04:08
Breath Volume		1.765	04:08
Air Blank		0.000	04:09
Air Blank		0.000	04:11
Subject Sample	# 2	0.227	04:11
Breath Volume		3.097	04:11
Air Blank		0.000	04:12
Control Test		0.082	04:12
Air Blank		0.000	04:13
Diagnostics Check		OK	04:13

Data Download Date/Time: 02/13/2018 17:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SULLIVAN,T
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11117080A3
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: GUZMAN-LOPEZ

DATE/TIME OF TEST: 01/16/2018 00:35:50

Diagnostics Check		OK	00:29
Air Blank		0.000	00:29
Control Test		0.081	00:29
Air Blank		0.000	00:30
Subject Sample	# 1	0.159	00:31
Breath Volume		1.277	00:31
Air Blank		0.000	00:31
Air Blank		0.000	00:33
Subject Sample	# 2	0.164	00:34
Breath Volume		1.792	00:34
Air Blank		0.000	00:34
Control Test		0.080	00:35
Air Blank		0.000	00:35
Diagnostics Check		OK	00:35

Data Download Date/Time: 02/13/2018 17:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKIPPER,ROGER,G
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEE,J 689
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11117080A3
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 12:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:04

SUBJECT'S LAST NAME: CARO VIAMONTES

DATE/TIME OF TEST: 01/16/2018 14:32:22

Diagnostics Check		OK	14:26
Air Blank		0.000	14:26
Control Test		0.081	14:26
Air Blank		0.000	14:27
Subject Sample	# 1	0.000	14:28
Breath Volume		2.792	14:28
Air Blank		0.000	14:28
Air Blank		0.000	14:30
Subject Sample	# 2	0.000	14:31
Breath Volume		2.601	14:31
Air Blank		0.000	14:31
Control Test		0.081	14:31
Air Blank		0.000	14:32
Diagnostics Check		OK	14:32

Data Download Date/Time: 02/13/2018 17:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRANKLIN,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11117080A3
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: GOODMAN

DATE/TIME OF TEST: 01/17/2018 00:54:28

Diagnostics Check		OK	00:51	
Air Blank		0.000	00:52	
Control Test		0.082	00:52	
Air Blank		0.000	00:52	
Subject Sample	# 1		00:53	Subject Test Refused
Breath Volume		0.000	00:53	
Air Blank		0.000	00:53	
Control Test		0.082	00:53	
Air Blank		0.000	00:54	
Diagnostics Check		OK	00:54	

Data Download Date/Time: 02/13/2018 17:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: EDGE,S
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11117080A3
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 01/18/2018 22:55:13

Diagnostics Check		OK	22:48
Air Blank		0.000	22:49
Control Test		0.083	22:49
Air Blank		0.000	22:50
Subject Sample	# 1	0.128	22:50
Breath Volume		1.800	22:50
Air Blank		0.000	22:51
Air Blank		0.000	22:53
Subject Sample	# 2	0.128	22:53
Breath Volume		1.851	22:53
Air Blank		0.000	22:54
Control Test		0.082	22:54
Air Blank		0.000	22:55
Diagnostics Check		OK	22:55

Data Download Date/Time: 02/13/2018 17:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11117080A3
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: BRITTON

DATE/TIME OF TEST: 01/19/2018 03:38:22

Diagnostics Check		OK	03:31
Air Blank		0.000	03:32
Control Test		0.082	03:32
Air Blank		0.000	03:33
Subject Sample	# 1	0.133	03:33
Breath Volume		3.261	03:33
Air Blank		0.000	03:34
Air Blank		0.000	03:36
Subject Sample	# 2	0.121	03:36
Breath Volume		2.000	03:36
Air Blank		0.000	03:37
Control Test		0.080	03:37
Air Blank		0.000	03:38
Diagnostics Check		OK	03:38

Data Download Date/Time: 02/13/2018 17:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11117080A3
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: CULLIFER

DATE/TIME OF TEST: 01/19/2018 23:09:30

Diagnostics Check		OK	23:03
Air Blank		0.000	23:03
Control Test		0.082	23:03
Air Blank		0.000	23:04
Subject Sample	# 1	0.000	23:05
Breath Volume		2.375	23:05
Air Blank		0.000	23:05
Air Blank		0.000	23:07
Subject Sample	# 2	0.000	23:08
Breath Volume		1.613	23:08
Air Blank		0.000	23:08
Control Test		0.081	23:08
Air Blank		0.000	23:09
Diagnostics Check		OK	23:09

Data Download Date/Time: 02/13/2018 17:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DEJESUS,V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11117080A3
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 01/20/2018 02:07:00

Diagnostics Check		OK	02:00
Air Blank		0.000	02:00
Control Test		0.082	02:01
Air Blank		0.000	02:01
Subject Sample	# 1	0.153	02:02
Breath Volume		1.371	02:02
Air Blank		0.000	02:02
Air Blank		0.000	02:04
Subject Sample	# 2	0.141	02:05
Breath Volume		2.054	02:05
Air Blank		0.000	02:06
Control Test		0.083	02:06
Air Blank		0.000	02:06
Diagnostics Check		OK	02:07

Data Download Date/Time: 02/13/2018 17:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11117080A3
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: MCGEE

DATE/TIME OF TEST: 01/20/2018 02:46:40

Diagnostics Check		OK	02:39
Air Blank		0.000	02:40
Control Test		0.082	02:40
Air Blank		0.000	02:40
Subject Sample	# 1	0.169	02:41
Breath Volume		2.179	02:41
Air Blank		0.000	02:42
Air Blank		0.000	02:44
Subject Sample	# 2	0.186	02:45
Breath Volume		3.164	02:45
Air Blank		0.000	02:45
Control Test		0.082	02:46
Air Blank		0.000	02:46
Diagnostics Check		OK	02:46

Data Download Date/Time: 02/13/2018 17:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRANKLIN,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11117080A3
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: CRUZ PADRON

DATE/TIME OF TEST: 01/20/2018 20:38:22

Diagnostics Check		OK	20:32
Air Blank		0.000	20:32
Control Test		0.082	20:32
Air Blank		0.000	20:33
Subject Sample	# 1	0.185	20:33
Breath Volume		1.992	20:33
Air Blank		0.000	20:34
Air Blank		0.000	20:36
Subject Sample	# 2	0.189	20:36
Breath Volume		2.011	20:36
Air Blank		0.000	20:37
Control Test		0.081	20:37
Air Blank		0.000	20:38
Diagnostics Check		OK	20:38

Data Download Date/Time: 02/13/2018 17:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11117080A3
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: WEST

DATE/TIME OF TEST: 01/21/2018 04:00:07

Diagnostics Check		OK	03:53
Air Blank		0.000	03:54
Control Test		0.081	03:54
Air Blank		0.000	03:54
Subject Sample	# 1	0.158	03:55
Breath Volume		3.074	03:55
Air Blank		0.000	03:56
Air Blank		0.000	03:57
Subject Sample	# 2	0.163	03:58
Breath Volume		3.386	03:58
Air Blank		0.000	03:59
Control Test		0.081	03:59
Air Blank		0.000	04:00
Diagnostics Check		OK	04:00

Data Download Date/Time: 02/13/2018 17:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRENNAN,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11117080A3
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 01/21/2018 04:28:27

Diagnostics Check		OK	04:22
Air Blank		0.000	04:22
Control Test		0.081	04:22
Air Blank		0.000	04:23
Subject Sample	# 1	0.089	04:24
Breath Volume		3.570	04:24
Air Blank		0.000	04:24
Air Blank		0.000	04:26
Subject Sample	# 2	0.090	04:26
Breath Volume		1.246	04:26
Air Blank		0.000	04:27
Control Test		0.081	04:27
Air Blank		0.000	04:28
Diagnostics Check		OK	04:28

Data Download Date/Time: 02/13/2018 17:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11117080A3
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: MADURO JR

DATE/TIME OF TEST: 01/22/2018 01:07:41

Diagnostics Check		OK	01:00
Air Blank		0.000	01:01
Control Test		0.082	01:01
Air Blank		0.000	01:02
Subject Sample	# 1	0.128	01:02
Breath Volume		4.058	01:02
Air Blank		0.000	01:03
Air Blank		0.000	01:05
Subject Sample	# 2	0.126	01:06
Breath Volume		3.718	01:06
Air Blank		0.000	01:06
Control Test		0.080	01:07
Air Blank		0.000	01:07
Diagnostics Check		OK	01:07

Data Download Date/Time: 02/13/2018 17:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS,RONALD,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARFIELD,C
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 01/22/2018 04:40:02

Diagnostics Check		OK	04:33
Air Blank		0.000	04:34
Control Test		0.080	04:34
Air Blank		0.000	04:34
Subject Sample	# 1	0.127	04:35
Breath Volume		2.468	04:35
Air Blank		0.000	04:36
Air Blank		0.000	04:37
Subject Sample	# 2	0.120	04:38
Breath Volume		2.210	04:38
Air Blank		0.000	04:39
Control Test		0.080	04:39
Air Blank		0.000	04:39
Diagnostics Check		OK	04:40

Data Download Date/Time: 02/13/2018 17:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: CONTEH

DATE/TIME OF TEST: 01/23/2018 03:55:56

Diagnostics Check		OK	03:49
Air Blank		0.000	03:49
Control Test		0.080	03:50
Air Blank		0.000	03:50
Subject Sample	# 1	0.171	03:51
Breath Volume		2.738	03:51
Air Blank		0.000	03:51
Air Blank		0.000	03:53
Subject Sample	# 2	0.169	03:54
Breath Volume		3.589	03:54
Air Blank		0.000	03:55
Control Test		0.079	03:55
Air Blank		0.000	03:55
Diagnostics Check		OK	03:55

Data Download Date/Time: 02/13/2018 17:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKIPPER,ROGER,G
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TEST,TEST
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: SYS CHECK
ARREST TIME: 12:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:00

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 01/23/2018 13:47:39

Diagnostics Check		OK	13:41
Air Blank		0.000	13:42
Control Test		0.079	13:42
Air Blank		0.000	13:43
Subject Sample	# 1	0.000	13:43
Breath Volume		1.414	13:43
Air Blank		0.000	13:43
Air Blank		0.000	13:45
Subject Sample	# 2	0.000	13:46
Breath Volume		1.257	13:46
Air Blank		0.000	13:46
Control Test		0.079	13:47
Air Blank		0.000	13:47
Diagnostics Check		OK	13:47

Data Download Date/Time: 02/13/2018 17:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAINTILIEN,BOB,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: SAINTILIEN,BOB
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 12/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: BERCEY

DATE/TIME OF TEST: 01/17/2018 04:15:25

Diagnostics Check		OK	04:04
Air Blank		0.000	04:04
Control Test		0.081	04:05
Air Blank		0.000	04:05
Subject Sample	# 1	0.170	04:07
Breath Volume		1.792	04:07
Air Blank		0.000	04:07
Air Blank		0.000	04:09
Subject Sample	# 2	0.144	04:10
Breath Volume		3.246	04:10
Air Blank		0.000	04:11
Air Blank		0.000	04:13
Subject Sample	# 3	0.157	04:13
Breath Volume		2.585	04:13
Air Blank		0.000	04:14
Control Test		0.081	04:14
Air Blank		0.000	04:15
Diagnostics Check		OK	04:15

Data Download Date/Time: 01/26/2018 10:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,WILLIAM,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GREEN,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 20:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: NORDYKE

DATE/TIME OF TEST: 01/09/2018 20:53:35

Diagnostics Check		OK	20:51	
Air Blank		0.000	20:51	
Control Test		0.078	20:51	
Air Blank		0.000	20:52	
Subject Sample	# 1		20:53	RFI Detect
Breath Volume		0.000	20:53	
Air Blank			20:53	RFI Detect
Air Blank			20:53	RFI Detect

Data Download Date/Time: 01/30/2018 13:53

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,WILLIAM,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GREEN,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 20:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: NORDYKE

DATE/TIME OF TEST: 01/09/2018 20:57:39

Diagnostics Check	OK	20:57	
Air Blank		20:57	RFI Detect
Air Blank		20:57	RFI Detect

Data Download Date/Time: 01/30/2018 13:53

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,WILLIAM,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GREEN,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 20:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: NORDYKE

DATE/TIME OF TEST: 01/09/2018 21:00:41

Diagnostics Check	OK	20:59	
Air Blank	0.000	20:59	
Control Test	0.078	21:00	Interferent Detect
Air Blank	0.000	21:00	

Data Download Date/Time: 01/30/2018 13:53

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,WILLIAM,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GREEN,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 20:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: NORDYKE

DATE/TIME OF TEST: 01/09/2018 21:10:39

Diagnostics Check		OK	21:03
Air Blank		0.000	21:03
Control Test		0.079	21:04
Air Blank		0.000	21:04
Subject Sample	# 1	0.183	21:06
Breath Volume		1.824	21:06
Air Blank		0.000	21:06
Air Blank		0.000	21:08
Subject Sample	# 2	0.187	21:09
Breath Volume		1.718	21:09
Air Blank		0.000	21:09
Control Test		0.078	21:10
Air Blank		0.000	21:10
Diagnostics Check		OK	21:10

Data Download Date/Time: 01/30/2018 13:53

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,WILLIAM,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GREEN,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 01/10/2018 23:47:14

Diagnostics Check		OK	23:40
Air Blank		0.000	23:41
Control Test		0.078	23:41
Air Blank		0.000	23:42
Subject Sample	# 1	0.094	23:42
Breath Volume		2.523	23:42
Air Blank		0.000	23:43
Air Blank		0.000	23:45
Subject Sample	# 2	0.087	23:45
Breath Volume		2.476	23:45
Air Blank		0.000	23:46
Control Test		0.078	23:46
Air Blank		0.000	23:47
Diagnostics Check		OK	23:47

Data Download Date/Time: 01/30/2018 13:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,WILLIAM,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GREEN,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 19:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 01/11/2018 20:15:24

Diagnostics Check		OK	20:11	
Air Blank		0.000	20:11	
Control Test		0.077	20:11	
Air Blank		0.000	20:12	
Subject Sample	# 1		20:14	Subject Test Refused
Breath Volume		0.000	20:14	
Air Blank		0.000	20:14	
Control Test		0.078	20:14	
Air Blank		0.000	20:15	
Diagnostics Check		OK	20:15	

Data Download Date/Time: 01/30/2018 13:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COFFEE,JASON,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: COFFEE,JASON
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: MEADOWS

DATE/TIME OF TEST: 01/16/2018 02:53:06

Diagnostics Check	OK	02:52	
Air Blank		02:52	RFI Detect
Air Blank	0.000	02:53	

Data Download Date/Time: 01/30/2018 13:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COFFEE,JASON,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: COFFEE,JASON
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: MEADOWS

DATE/TIME OF TEST: 01/16/2018 03:01:28

Diagnostics Check		OK	02:57	
Air Blank		0.000	02:57	
Control Test		0.079	02:58	
Air Blank		0.000	02:58	
Subject Sample	# 1	0.107	02:59	
Breath Volume		2.128	02:59	
Air Blank			03:00	RFI Detect
Air Blank		0.000	03:00	
Control Test		0.078	03:00	
Air Blank		0.000	03:01	
Diagnostics Check		OK	03:01	

Data Download Date/Time: 01/30/2018 13:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COFFEE,JASON,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: COFFEE,JASON
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: MEADOWS

DATE/TIME OF TEST: 01/16/2018 03:11:35

Diagnostics Check		OK	03:04
Air Blank		0.000	03:04
Control Test		0.079	03:05
Air Blank		0.000	03:05
Subject Sample	# 1	0.106	03:06
Breath Volume		1.710	03:06
Air Blank		0.000	03:07
Air Blank		0.000	03:08
Subject Sample	# 2	0.101	03:10
Breath Volume		1.910	03:10
Air Blank		0.000	03:10
Control Test		0.078	03:10
Air Blank		0.000	03:11
Diagnostics Check		OK	03:11

Data Download Date/Time: 01/30/2018 13:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,WILLIAM,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GREEN,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: DOLPHIN

DATE/TIME OF TEST: 01/21/2018 02:09:25

Diagnostics Check		OK	02:02
Air Blank		0.000	02:02
Control Test		0.079	02:03
Air Blank		0.000	02:03
Subject Sample	# 1	0.170	02:04
Breath Volume		2.167	02:04
Air Blank		0.000	02:05
Air Blank		0.000	02:06
Subject Sample	# 2	0.168	02:07
Breath Volume		2.109	02:07
Air Blank		0.000	02:08
Control Test		0.078	02:08
Air Blank		0.000	02:09
Diagnostics Check		OK	02:09

Data Download Date/Time: 01/30/2018 13:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,WILLIAM,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GREEN,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: CHAPMAN

DATE/TIME OF TEST: 01/25/2018 00:19:54

Diagnostics Check		OK	00:13
Air Blank		0.000	00:13
Control Test		0.079	00:14
Air Blank		0.000	00:14
Subject Sample	# 1	0.178	00:15
Breath Volume		2.468	00:15
Air Blank		0.000	00:15
Air Blank		0.000	00:17
Subject Sample	# 2	0.181	00:18
Breath Volume		2.812	00:18
Air Blank		0.000	00:18
Control Test		0.079	00:19
Air Blank		0.000	00:19
Diagnostics Check		OK	00:19

Data Download Date/Time: 01/30/2018 13:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPOS,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CAMPOS,MICHAEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 01/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: RODRIGUEZ VICENTE

DATE/TIME OF TEST: 01/17/2018 23:15:20

Diagnostics Check		OK	23:06
Air Blank		0.000	23:07
Control Test		0.082	23:07
Air Blank		0.000	23:08
Subject Sample	# 1	0.161	23:09
Breath Volume		1.671	23:09
Air Blank		0.000	23:11
Air Blank		0.000	23:12
Subject Sample	# 2	0.155	23:13
Breath Volume		1.445	23:13
Air Blank		0.000	23:14
Control Test		0.075	23:14
Air Blank		0.000	23:15
Diagnostics Check		OK	23:15

Data Download Date/Time: 02/01/2018 19:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.0.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,DANIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LYONS,DANIEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 01/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: SIERRA CRUZ

DATE/TIME OF TEST: 01/11/2018 03:04:13

Diagnostics Check		OK	02:57
Air Blank		0.000	02:58
Control Test		0.078	02:58
Air Blank		0.000	02:59
Subject Sample	# 1	0.220	02:59
Breath Volume		1.253	02:59
Air Blank		0.000	03:00
Air Blank		0.000	03:02
Subject Sample	# 2	0.224	03:02
Breath Volume		1.789	03:02
Air Blank		0.000	03:03
Control Test		0.078	03:03
Air Blank		0.000	03:04
Diagnostics Check		OK	03:04

Data Download Date/Time: 02/01/2018 19:40

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.0.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,DANIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: MCCONNELL,CHARLES
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 01/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: DOOD

DATE/TIME OF TEST: 01/16/2018 23:39:32

Diagnostics Check		OK	23:34	
Air Blank		0.000	23:34	
Control Test		0.081	23:35	
Air Blank		0.000	23:35	
Subject Sample	# 1	0.076	23:35	
Breath Volume		2.699	23:35	
Air Blank		0.000	23:36	
Air Blank		0.000	23:38	
Subject Sample	# 2	0.075	23:38	
Breath Volume		3.929	23:38	
Air Blank			23:39	RFI Detect
Air Blank			23:39	RFI Detect

Data Download Date/Time: 02/01/2018 19:40

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.0.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,DANIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: MCCONNELL,CHARLES
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 01/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: DOOD

DATE/TIME OF TEST: 01/16/2018 23:47:56

Diagnostics Check		OK	23:42
Air Blank		0.000	23:42
Control Test		0.080	23:42
Air Blank		0.000	23:43
Subject Sample	# 1	0.070	23:43
Breath Volume		2.390	23:43
Air Blank		0.000	23:44
Air Blank		0.000	23:46
Subject Sample	# 2	0.068	23:46
Breath Volume		3.250	23:46
Air Blank		0.000	23:47
Control Test		0.079	23:47
Air Blank		0.000	23:47
Diagnostics Check		OK	23:47

Data Download Date/Time: 02/01/2018 19:40

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.0.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,DANIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LYONS,DANIEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 01/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: RYAN

DATE/TIME OF TEST: 01/17/2018 00:53:47

Diagnostics Check		OK	00:50	
Air Blank		0.000	00:50	
Control Test		0.080	00:51	
Air Blank		0.000	00:51	
Subject Sample	# 1		00:52	RFI Detect
Breath Volume		0.273	00:52	
Air Blank		0.000	00:52	
Control Test		0.080	00:53	
Air Blank			00:53	RFI Detect
Air Blank		0.000	00:53	

Data Download Date/Time: 02/01/2018 19:40

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.0.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,DANIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LYONS,DANIEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 01/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: RYAN

DATE/TIME OF TEST: 01/17/2018 01:08:00

Diagnostics Check		OK	00:56	
Air Blank		0.000	00:57	
Control Test		0.080	00:57	
Air Blank		0.000	00:57	
Subject Sample	# 1	0.000	01:01	Volume Not Met
Breath Volume		0.238	01:01	
Air Blank		0.000	01:01	
Air Blank		0.000	01:03	
Subject Sample	# 2	0.000	01:06	Volume Not Met
Breath Volume		0.703	01:06	
Air Blank		0.000	01:07	
Control Test		0.078	01:07	
Air Blank		0.000	01:07	
Diagnostics Check		OK	01:08	

Data Download Date/Time: 02/01/2018 19:40

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.0.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS',DANIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LYONS,DANIEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 01/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 18:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:37

SUBJECT'S LAST NAME: CORNETT

DATE/TIME OF TEST: 01/19/2018 19:04:09

Diagnostics Check		OK	18:58
Air Blank		0.000	18:58
Control Test		0.081	18:58
Air Blank		0.000	18:59
Subject Sample	# 1	0.405	18:59
Breath Volume		1.804	18:59
Air Blank		0.000	19:00
Air Blank		0.000	19:02
Subject Sample	# 2	0.401	19:02
Breath Volume		1.828	19:02
Air Blank		0.000	19:03
Control Test		0.080	19:03
Air Blank		0.000	19:04
Diagnostics Check		OK	19:04

Data Download Date/Time: 02/01/2018 19:40

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.0.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,DANIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: DUNN,COLIN
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 01/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:46

SUBJECT'S LAST NAME: SMILIE

DATE/TIME OF TEST: 01/21/2018 22:14:32

Diagnostics Check		OK	22:06
Air Blank		0.000	22:07
Control Test		0.080	22:07
Air Blank		0.000	22:08
Subject Sample	# 1	0.207	22:09
Breath Volume		1.667	22:09
Air Blank		0.000	22:10
Air Blank		0.000	22:12
Subject Sample	# 2	0.205	22:12
Breath Volume		2.199	22:12
Air Blank		0.000	22:13
Control Test		0.077	22:13
Air Blank		0.000	22:14
Diagnostics Check		OK	22:14

Data Download Date/Time: 02/01/2018 19:40

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.0.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,DANIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LYONS,DANIEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 01/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: ANDUJAR

DATE/TIME OF TEST: 01/26/2018 02:21:44

Diagnostics Check		OK	02:13
Air Blank		0.000	02:13
Control Test		0.080	02:14
Air Blank		0.000	02:14
Subject Sample	# 1	0.156	02:15
Breath Volume		1.503	02:15
Air Blank		0.000	02:16
Air Blank		0.000	02:17
Subject Sample	# 2	0.161	02:20
Breath Volume		1.386	02:20
Air Blank		0.000	02:20
Control Test		0.081	02:21
Air Blank		0.000	02:21
Diagnostics Check		OK	02:21

Data Download Date/Time: 02/01/2018 19:40

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULDERRIG,SEAN,F
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: MULDERRIG,SEAN
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: FORSYTHE

DATE/TIME OF TEST: 01/01/2018 22:56:08

Diagnostics Check		OK	22:53	
Air Blank		0.000	22:53	
Control Test		0.080	22:54	
Air Blank		0.000	22:54	
Subject Sample	# 1		22:54	Subject Test Refused
Breath Volume		0.000	22:54	
Air Blank		0.000	22:55	
Control Test		0.081	22:55	
Air Blank		0.000	22:56	
Diagnostics Check		OK	22:56	

Data Download Date/Time: 01/23/2018 09:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTLE,CORY,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SUTTLE,CORY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: POZERYCKI

DATE/TIME OF TEST: 01/05/2018 02:55:02

Diagnostics Check		OK	02:39	
Air Blank		0.000	02:39	
Control Test		0.080	02:40	
Air Blank		0.000	02:40	
Subject Sample	# 1	0.140	02:43	
Breath Volume		1.679	02:43	
Air Blank		0.000	02:44	
Air Blank		0.000	02:45	
Subject Sample	# 2	0.113	02:49	Volume Not Met
Breath Volume		0.277	02:49	
Air Blank		0.000	02:49	
Air Blank		0.000	02:51	
Subject Sample	# 3	0.141	02:53	
Breath Volume		1.550	02:53	
Air Blank		0.000	02:54	
Control Test		0.080	02:54	
Air Blank		0.000	02:54	
Diagnostics Check		OK	02:55	

Data Download Date/Time: 01/23/2018 09:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTLE,CORY,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SUTTLE,CORY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: CATES

DATE/TIME OF TEST: 01/05/2018 23:05:58

Diagnostics Check		OK	23:02	
Air Blank		0.000	23:03	
Control Test		0.081	23:03	
Air Blank		0.000	23:04	
Subject Sample	# 1		23:04	Subject Test Refused
Breath Volume		0.000	23:04	
Air Blank		0.000	23:05	
Control Test		0.081	23:05	
Air Blank		0.000	23:05	
Diagnostics Check		OK	23:05	

Data Download Date/Time: 01/23/2018 09:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTLE,CORY,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SUTTLE,CORY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: MACDONALD

DATE/TIME OF TEST: 01/06/2018 21:43:31

Diagnostics Check		OK	21:37
Air Blank		0.000	21:37
Control Test		0.081	21:37
Air Blank		0.000	21:38
Subject Sample	# 1	0.207	21:38
Breath Volume		2.589	21:38
Air Blank		0.000	21:39
Air Blank		0.000	21:41
Subject Sample	# 2	0.213	21:41
Breath Volume		2.808	21:41
Air Blank		0.000	21:42
Control Test		0.079	21:42
Air Blank		0.000	21:43
Diagnostics Check		OK	21:43

Data Download Date/Time: 01/23/2018 09:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTLE,CORY,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SUTTLE,CORY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: TALLEY

DATE/TIME OF TEST: 01/07/2018 02:53:01

Diagnostics Check		OK	02:46
Air Blank		0.000	02:47
Control Test		0.080	02:47
Air Blank		0.000	02:48
Subject Sample	# 1	0.000	02:48
Breath Volume		4.277	02:48
Air Blank		0.000	02:49
Air Blank		0.000	02:51
Subject Sample	# 2	0.000	02:51
Breath Volume		4.265	02:51
Air Blank		0.000	02:52
Control Test		0.080	02:52
Air Blank		0.000	02:52
Diagnostics Check		OK	02:53

Data Download Date/Time: 01/23/2018 09:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULDERRIG,SEAN,F
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: MULDERRIG,SEAN
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 06:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:01

SUBJECT'S LAST NAME: PAGAN

DATE/TIME OF TEST: 01/07/2018 07:34:09

Diagnostics Check		OK	07:22	
Air Blank		0.000	07:23	
Control Test		0.080	07:23	
Air Blank		0.000	07:24	
Subject Sample	# 1		07:27	No Sample Provided
Breath Volume		0.000	07:27	
Air Blank		0.000	07:27	
Air Blank		0.000	07:29	
Subject Sample	# 2		07:32	No Sample Provided
Breath Volume		0.000	07:32	
Air Blank		0.000	07:33	
Control Test		0.080	07:33	
Air Blank		0.000	07:34	
Diagnostics Check		OK	07:34	

Data Download Date/Time: 01/23/2018 09:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULDERRIG,SEAN,F
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: MULDERRIG,SEAN
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: MARISCAL-MARTINEZ

DATE/TIME OF TEST: 01/10/2018 23:49:34

Diagnostics Check		OK	23:40
Air Blank		0.000	23:41
Control Test		0.080	23:41
Air Blank		0.000	23:41
Subject Sample	# 1	0.133	23:43
Breath Volume		2.289	23:43
Air Blank		0.000	23:44
Air Blank		0.000	23:46
Subject Sample	# 2	0.126	23:47
Breath Volume		1.730	23:47
Air Blank		0.000	23:48
Control Test		0.079	23:48
Air Blank		0.000	23:49
Diagnostics Check		OK	23:49

Data Download Date/Time: 01/23/2018 09:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTLE,CORY,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SUTTLE,CORY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 01/12/2018 00:21:39

Diagnostics Check		OK	00:15
Air Blank		0.000	00:15
Control Test		0.079	00:16
Air Blank		0.000	00:16
Subject Sample	# 1	0.130	00:17
Breath Volume		1.718	00:17
Air Blank		0.000	00:17
Air Blank		0.000	00:19
Subject Sample	# 2	0.129	00:20
Breath Volume		1.500	00:20
Air Blank		0.000	00:20
Control Test		0.078	00:21
Air Blank		0.000	00:21
Diagnostics Check		OK	00:21

Data Download Date/Time: 01/23/2018 09:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTLE,CORY,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SUTTLE,CORY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: RAY

DATE/TIME OF TEST: 01/12/2018 23:47:08

Diagnostics Check		OK	23:44	
Air Blank		0.000	23:44	
Control Test		0.080	23:44	
Air Blank		0.000	23:45	
Subject Sample	# 1		23:45	Subject Test Refused
Breath Volume		0.000	23:45	
Air Blank		0.000	23:46	
Control Test		0.079	23:46	
Air Blank		0.000	23:47	
Diagnostics Check		OK	23:47	

Data Download Date/Time: 01/23/2018 09:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTLE,CORY,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SUTTLE,CORY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 01/14/2018 03:37:19

Diagnostics Check		OK	03:30
Air Blank		0.000	03:31
Control Test		0.081	03:31
Air Blank		0.000	03:32
Subject Sample	# 1	0.195	03:32
Breath Volume		1.949	03:32
Air Blank		0.000	03:33
Air Blank		0.000	03:35
Subject Sample	# 2	0.193	03:35
Breath Volume		1.769	03:35
Air Blank		0.000	03:36
Control Test		0.081	03:36
Air Blank		0.000	03:37
Diagnostics Check		OK	03:37

Data Download Date/Time: 01/23/2018 09:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTLE,CORY,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SUTTLE,CORY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: IDELBURG

DATE/TIME OF TEST: 01/20/2018 23:49:58

Diagnostics Check		OK	23:43
Air Blank		0.000	23:43
Control Test		0.080	23:44
Air Blank		0.000	23:44
Subject Sample	# 1	0.123	23:45
Breath Volume		1.656	23:45
Air Blank		0.000	23:46
Air Blank		0.000	23:47
Subject Sample	# 2	0.132	23:48
Breath Volume		1.238	23:48
Air Blank		0.000	23:49
Control Test		0.079	23:49
Air Blank		0.000	23:49
Diagnostics Check		OK	23:49

Data Download Date/Time: 01/23/2018 09:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEEKMAN,JACOB,C
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: BEEKMAN,JACOB
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 01/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: SEIJAS-PULIDO

DATE/TIME OF TEST: 01/14/2018 02:46:17

Diagnostics Check		OK	02:42	
Air Blank		0.000	02:43	
Control Test		0.080	02:43	
Air Blank		0.000	02:43	
Subject Sample	# 1		02:44	Slope Not Met
Breath Volume		3.265	02:44	
Air Blank		0.000	02:45	
Control Test		0.080	02:45	
Air Blank		0.000	02:46	
Diagnostics Check		OK	02:46	

Data Download Date/Time: 02/01/2018 19:10

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEEKMAN,JACOB,C
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: BEEKMAN,JACOB
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 01/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: SEIJAS-PULIDO

DATE/TIME OF TEST: 01/14/2018 02:55:09

Diagnostics Check		OK	02:49	
Air Blank		0.000	02:50	
Control Test		0.080	02:50	
Air Blank		0.000	02:50	
Subject Sample	# 1	0.150	02:52	
Breath Volume		3.082	02:52	
Air Blank		0.000	02:52	
Air Blank			02:54	Ambient Fail
Air Blank		0.000	02:55	

Data Download Date/Time: 02/01/2018 19:10

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDAWAY,BRANDON,R
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: HARDAWAY,BRANDON
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 12/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: NOEL

DATE/TIME OF TEST: 01/13/2018 02:29:47

Diagnostics Check		OK	02:26	
Air Blank		0.000	02:27	
Control Test		0.079	02:27	
Air Blank		0.000	02:28	
Subject Sample	# 1		02:28	Subject Test Refused
Breath Volume		0.000	02:28	
Air Blank		0.000	02:28	
Control Test		0.079	02:29	
Air Blank		0.000	02:29	
Diagnostics Check		OK	02:29	

Data Download Date/Time: 01/29/2018 10:15

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNLAP,GABRIEL,C
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: WALKER,LANDON
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 12/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 01/13/2018 23:42:15

Diagnostics Check		OK	23:38	
Air Blank		0.000	23:39	
Control Test		0.079	23:39	
Air Blank		0.000	23:39	
Subject Sample	# 1		23:41	RFI Detect
Breath Volume		0.000	23:41	
Air Blank			23:41	RFI Detect
Air Blank		0.000	23:42	

Data Download Date/Time: 01/29/2018 10:15

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNLAP,GABRIEL,C
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: WALKER,LANDON
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 12/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: HERNANDEZ RIVERA

DATE/TIME OF TEST: 01/13/2018 23:55:22

Diagnostics Check		OK	23:45	
Air Blank		0.000	23:45	
Control Test		0.079	23:46	
Air Blank		0.000	23:46	
Subject Sample	# 1		23:49	No Sample Provided
Breath Volume		0.000	23:49	
Air Blank		0.000	23:50	
Air Blank		0.000	23:52	
Subject Sample	# 2		23:54	Sequence Aborted
Breath Volume		0.000	23:54	
Air Blank		0.000	23:55	

Data Download Date/Time: 01/29/2018 10:15

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER,TYLER,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FOWLER,TYLER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 01/01/2018 02:05:51

Diagnostics Check		OK	01:58
Air Blank		0.000	01:58
Control Test		0.078	01:59
Air Blank		0.000	01:59
Subject Sample	# 1	0.184	02:01
Breath Volume		2.562	02:01
Air Blank		0.000	02:02
Air Blank		0.000	02:03
Subject Sample	# 2	0.192	02:04
Breath Volume		1.906	02:04
Air Blank		0.000	02:04
Control Test		0.078	02:05
Air Blank		0.000	02:05
Diagnostics Check		OK	02:05

Data Download Date/Time: 01/31/2018 07:48

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLINT,SERGEANT,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FLINT,SERGEANT
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:49

SUBJECT'S LAST NAME: TAPIA

DATE/TIME OF TEST: 01/12/2018 19:16:36

Diagnostics Check		OK	19:10
Air Blank		0.000	19:11
Control Test		0.078	19:11
Air Blank		0.000	19:11
Subject Sample	# 1	0.000	19:12
Breath Volume		1.214	19:12
Air Blank		0.000	19:12
Air Blank		0.000	19:14
Subject Sample	# 2	0.000	19:15
Breath Volume		1.140	19:15
Air Blank		0.000	19:15
Control Test		0.079	19:16
Air Blank		0.000	19:16
Diagnostics Check		OK	19:16

Data Download Date/Time: 01/31/2018 07:48

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER,TYLER,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FOWLER,TYLER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: KOTWICKI

DATE/TIME OF TEST: 01/13/2018 03:52:50

Diagnostics Check		OK	03:46
Air Blank		0.000	03:46
Control Test		0.078	03:47
Air Blank		0.000	03:47
Subject Sample	# 1	0.283	03:48
Breath Volume		4.441	03:48
Air Blank		0.000	03:48
Air Blank		0.000	03:50
Subject Sample	# 2	0.275	03:51
Breath Volume		3.402	03:51
Air Blank		0.000	03:51
Control Test		0.078	03:52
Air Blank		0.000	03:52
Diagnostics Check		OK	03:52

Data Download Date/Time: 01/31/2018 07:48

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER,TYLER,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FOWLER,T
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:43

SUBJECT'S LAST NAME: SWAN

DATE/TIME OF TEST: 01/24/2018 23:12:03

Diagnostics Check		OK	23:03
Air Blank		0.000	23:04
Control Test		0.080	23:04
Air Blank		0.000	23:05
Subject Sample	# 1	0.090	23:05
Breath Volume		1.191	23:05
Air Blank		0.000	23:06
Air Blank		0.000	23:07
Subject Sample	# 2	0.088	23:10
Breath Volume		2.460	23:10
Air Blank		0.000	23:11
Control Test		0.078	23:11
Air Blank		0.000	23:11
Diagnostics Check		OK	23:12

Data Download Date/Time: 01/31/2018 07:48

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER,TYLER,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FOWLER,TYLER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: BONARO

DATE/TIME OF TEST: 01/26/2018 23:32:28

Diagnostics Check		OK	23:22
Air Blank		0.000	23:22
Control Test		0.079	23:23
Air Blank		0.000	23:23
Subject Sample	# 1	0.192	23:25
Breath Volume		1.625	23:25
Air Blank		0.000	23:26
Air Blank		0.000	23:28
Subject Sample	# 2	0.184	23:30
Breath Volume		1.671	23:30
Air Blank		0.000	23:31
Control Test		0.076	23:31
Air Blank		0.000	23:32
Diagnostics Check		OK	23:32

Data Download Date/Time: 01/31/2018 07:48

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER,TYLER,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FOWLER,TYLER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: WEBB JR

DATE/TIME OF TEST: 01/28/2018 00:02:49

Diagnostics Check		OK	23:56
Air Blank		0.000	23:56
Control Test		0.079	23:57
Air Blank		0.000	23:57
Subject Sample	# 2	0.145	23:58
Breath Volume		3.011	23:58
Air Blank		0.000	23:59
Air Blank		0.000	00:00
Subject Sample	# 1	0.154	00:01
Breath Volume		1.507	00:01
Air Blank		0.000	00:01
Control Test		0.078	00:02
Air Blank		0.000	00:02
Diagnostics Check		OK	00:02

Data Download Date/Time: 01/31/2018 07:48

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER,TYLER,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FOWLER,TYLER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: HOLMAN

DATE/TIME OF TEST: 01/31/2018 03:07:41

Diagnostics Check		OK	03:00
Air Blank		0.000	03:00
Control Test		0.081	03:01
Air Blank		0.000	03:01
Subject Sample	# 1	0.187	03:03
Breath Volume		2.304	03:03
Air Blank		0.000	03:03
Air Blank		0.000	03:05
Subject Sample	# 2	0.188	03:06
Breath Volume		1.992	03:06
Air Blank		0.000	03:06
Control Test		0.080	03:07
Air Blank		0.000	03:07
Diagnostics Check		OK	03:07

Data Download Date/Time: 01/31/2018 07:48

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER,TYLER,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FOWLER,TYLER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:46

SUBJECT'S LAST NAME: CARCATERRA

DATE/TIME OF TEST: 01/31/2018 22:12:20

Diagnostics Check		OK	22:06
Air Blank		0.000	22:06
Control Test		0.080	22:07
Air Blank		0.000	22:07
Subject Sample	# 1	0.085	22:07
Breath Volume		2.250	22:07
Air Blank		0.000	22:08
Air Blank		0.000	22:10
Subject Sample	# 2	0.087	22:10
Breath Volume		1.246	22:10
Air Blank		0.000	22:11
Control Test		0.079	22:11
Air Blank		0.000	22:12
Diagnostics Check		OK	22:12

Data Download Date/Time: 02/23/2018 07:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SHEINBERG,T
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: NEYLAND

DATE/TIME OF TEST: 01/01/2018 02:42:30

Diagnostics Check		OK	02:35
Air Blank		0.000	02:35
Control Test		0.083	02:36
Air Blank		0.000	02:36
Subject Sample	# 1	0.105	02:38
Breath Volume		1.179	02:38
Air Blank		0.000	02:38
Air Blank		0.000	02:40
Subject Sample	# 2	0.109	02:40
Breath Volume		1.355	02:40
Air Blank		0.000	02:41
Control Test		0.082	02:41
Air Blank		0.000	02:42
Diagnostics Check		OK	02:42

Data Download Date/Time: 01/03/2018 19:30

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 01/01/2018 04:58:33

Diagnostics Check		OK	04:55	
Air Blank		0.000	04:55	
Control Test		0.082	04:56	
Air Blank		0.000	04:56	
Subject Sample	# 1		04:57	Subject Test Refused
Breath Volume		0.000	04:57	
Air Blank		0.000	04:57	
Control Test		0.082	04:57	
Air Blank		0.000	04:58	
Diagnostics Check		OK	04:58	

Data Download Date/Time: 01/03/2018 19:30

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MACIAS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: MONTANEZ

DATE/TIME OF TEST: 01/03/2018 03:22:35

Diagnostics Check		OK	03:15
Air Blank		0.000	03:15
Control Test		0.083	03:16
Air Blank		0.000	03:16
Subject Sample	# 1	0.000	03:17
Breath Volume		1.886	03:17
Air Blank		0.000	03:18
Air Blank		0.000	03:20
Subject Sample	# 2	0.000	03:21
Breath Volume		1.800	03:21
Air Blank		0.000	03:21
Control Test		0.083	03:21
Air Blank		0.000	03:22
Diagnostics Check		OK	03:22

Data Download Date/Time: 01/03/2018 19:30

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS,RONALD,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DEJESUS,V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: PEDERSEN

DATE/TIME OF TEST: 01/03/2018 04:57:47

Diagnostics Check		OK	04:50
Air Blank		0.000	04:50
Control Test		0.083	04:51
Air Blank		0.000	04:51
Subject Sample	# 1	0.138	04:52
Breath Volume		2.363	04:52
Air Blank		0.000	04:53
Air Blank		0.000	04:55
Subject Sample	# 2	0.138	04:56
Breath Volume		2.054	04:56
Air Blank		0.000	04:56
Control Test		0.082	04:57
Air Blank		0.000	04:57
Diagnostics Check		OK	04:57

Data Download Date/Time: 01/03/2018 19:31

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,ANGELA,V
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: COVELL,C
ARRESTING OFFICER AGENCY: Macdill Air Force Ba
LAST AGENCY INSPECTION: 12/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:06

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 01/03/2018 07:32:54

Diagnostics Check		OK	07:26
Air Blank		0.000	07:27
Control Test		0.082	07:27
Air Blank		0.000	07:28
Subject Sample	# 1	0.122	07:28
Breath Volume		1.636	07:28
Air Blank		0.000	07:29
Air Blank		0.000	07:31
Subject Sample	# 2	0.119	07:31
Breath Volume		1.175	07:31
Air Blank		0.000	07:31
Control Test		0.082	07:32
Air Blank		0.000	07:32
Diagnostics Check		OK	07:32

Data Download Date/Time: 01/03/2018 19:31

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NIEUWENDAM,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 01/28/2018 01:10:39

Diagnostics Check		OK	01:05	
Air Blank		0.000	01:05	
Control Test		0.083	01:06	
Air Blank		0.000	01:06	
Subject Sample	# 1		01:09	Subject Test Refused
Breath Volume		0.000	01:09	
Air Blank		0.000	01:09	
Control Test		0.082	01:10	
Air Blank		0.000	01:10	
Diagnostics Check		OK	01:10	

Data Download Date/Time: 02/14/2018 09:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: YUN

DATE/TIME OF TEST: 01/28/2018 01:41:20

Diagnostics Check		OK	01:32	
Air Blank		0.000	01:33	
Control Test		0.083	01:33	
Air Blank		0.000	01:34	
Subject Sample	# 1		01:37	No Sample Provided
Breath Volume		0.000	01:37	
Air Blank		0.000	01:37	
Air Blank		0.000	01:39	
Subject Sample	# 2		01:39	Subject Test Refused
Breath Volume		0.000	01:39	
Air Blank		0.000	01:40	
Control Test		0.083	01:40	
Air Blank		0.000	01:41	
Diagnostics Check		OK	01:41	

Data Download Date/Time: 02/14/2018 09:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NIEUWENDAM,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 01/28/2018 04:11:23

Diagnostics Check		OK	04:04
Air Blank		0.000	04:05
Control Test		0.082	04:05
Air Blank		0.000	04:06
Subject Sample	# 1	0.171	04:07
Breath Volume		1.613	04:07
Air Blank		0.000	04:07
Air Blank		0.000	04:09
Subject Sample	# 2	0.166	04:09
Breath Volume		1.429	04:09
Air Blank		0.000	04:10
Control Test		0.082	04:10
Air Blank		0.000	04:11
Diagnostics Check		OK	04:11

Data Download Date/Time: 02/14/2018 09:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: EDGE,S
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: IBARRA

DATE/TIME OF TEST: 01/28/2018 05:09:08

Diagnostics Check		OK	05:02
Air Blank		0.000	05:03
Control Test		0.082	05:03
Air Blank		0.000	05:03
Subject Sample	# 1	0.155	05:04
Breath Volume		2.980	05:04
Air Blank		0.000	05:05
Air Blank		0.000	05:07
Subject Sample	# 2	0.151	05:07
Breath Volume		1.273	05:07
Air Blank		0.000	05:08
Control Test		0.082	05:08
Air Blank		0.000	05:09
Diagnostics Check		OK	05:09

Data Download Date/Time: 02/14/2018 09:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005146

INSTRUMENT REGISTERED TO: PATRICK AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OUVERSON,JESSE,T
BREATH TEST OPERATOR AGENCY: Patrick AFB PD
ARRESTING OFFICER: OUVERSON,JESSE
ARRESTING OFFICER AGENCY: Patrick AFB PD
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: SYS CHECK
ARREST TIME: 11:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:00

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 01/04/2018 13:35:22

Diagnostics Check	OK	13:34	
Air Blank	0.000	13:34	
Control Test	0.024	13:34	Control Outside Tolerance
Air Blank	0.000	13:35	

Data Download Date/Time: 01/24/2018 13:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005146

INSTRUMENT REGISTERED TO: PATRICK AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OUVERSON,JESSE,T
BREATH TEST OPERATOR AGENCY: Patrick AFB PD
ARRESTING OFFICER: OUVERSON,JESSE
ARRESTING OFFICER AGENCY: Patrick AFB PD
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: OTHER
ARREST TIME: 11:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:00

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 01/04/2018 13:38:04

Diagnostics Check	OK	13:36	
Air Blank	0.000	13:37	
Control Test	0.020	13:37	Control Outside Tolerance
Air Blank	0.000	13:38	

Data Download Date/Time: 01/24/2018 13:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005146

INSTRUMENT REGISTERED TO: PATRICK AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEST,TEST,T
BREATH TEST OPERATOR AGENCY: Patrick AFB PD
ARRESTING OFFICER: TEST,TEST
ARRESTING OFFICER AGENCY: Patrick AFB PD
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: ADMIN
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:00

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 01/24/2018 12:34:19

Diagnostics Check	OK	12:33	
Air Blank	0.000	12:33	
Control Test	0.007	12:33	Control Outside Tolerance
Air Blank	0.000	12:34	

Data Download Date/Time: 01/24/2018 13:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005146

INSTRUMENT REGISTERED TO: PATRICK AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OUVERSON,JESSE,T
BREATH TEST OPERATOR AGENCY: Patrick AFB PD
ARRESTING OFFICER: TEST,TEST
ARRESTING OFFICER AGENCY: Patrick AFB PD
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: ZERO TOL
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:00

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 01/24/2018 12:47:13

Diagnostics Check		OK	12:41	
Air Blank		0.000	12:41	
Control Test		0.081	12:42	
Air Blank		0.000	12:42	
Subject Sample	# 1	0.000	12:43	
Breath Volume		2.933	12:43	
Air Blank		0.000	12:43	
Air Blank		0.000	12:45	
Subject Sample	# 2		12:45	Subject Test Refused
Breath Volume		0.000	12:45	
Air Blank		0.000	12:46	
Control Test		0.081	12:46	
Air Blank		0.000	12:47	
Diagnostics Check		OK	12:47	

Data Download Date/Time: 01/24/2018 13:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COCHRANE,AARON,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: COCHRANE,AARON
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 12/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A2
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: DEPIETRO

DATE/TIME OF TEST: 01/01/2018 01:00:42

Diagnostics Check		OK	00:54
Air Blank		0.000	00:54
Control Test		0.082	00:54
Air Blank		0.000	00:55
Subject Sample	# 1	0.209	00:56
Breath Volume		2.714	00:56
Air Blank		0.000	00:56
Air Blank		0.000	00:58
Subject Sample	# 2	0.204	00:59
Breath Volume		3.492	00:59
Air Blank		0.000	00:59
Control Test		0.081	01:00
Air Blank		0.000	01:00
Diagnostics Check		OK	01:00

Data Download Date/Time: 01/15/2018 16:18

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STIMPSON,RACHEL,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MATHIS,JOHN
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 12/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A2
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: FRISCH

DATE/TIME OF TEST: 01/01/2018 23:22:12

Diagnostics Check		OK	23:09	
Air Blank		0.000	23:10	
Control Test		0.083	23:10	
Air Blank		0.000	23:11	
Subject Sample	# 1	0.129	23:14	Volume Not Met
Breath Volume		0.957	23:14	
Air Blank		0.000	23:15	
Air Blank		0.000	23:16	
Subject Sample	# 2	0.129	23:17	
Breath Volume		2.144	23:17	
Air Blank		0.000	23:18	
Air Blank		0.000	23:20	
Subject Sample	# 3	0.137	23:20	
Breath Volume		2.457	23:20	
Air Blank		0.000	23:21	
Control Test		0.078	23:21	
Air Blank		0.000	23:22	
Diagnostics Check		OK	23:22	

Data Download Date/Time: 01/15/2018 16:18

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIMA,PATRICIA,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MATHIS,J
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 12/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A2
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: LIMA

DATE/TIME OF TEST: 01/05/2018 23:16:48

Diagnostics Check		OK	23:10
Air Blank		0.000	23:11
Control Test		0.084	23:11
Air Blank		0.000	23:11
Subject Sample	# 1	0.210	23:12
Breath Volume		2.355	23:12
Air Blank		0.000	23:13
Air Blank		0.000	23:14
Subject Sample	# 2	0.215	23:15
Breath Volume		1.957	23:15
Air Blank		0.000	23:15
Control Test		0.081	23:16
Air Blank		0.000	23:16
Diagnostics Check		OK	23:16

Data Download Date/Time: 01/15/2018 16:18

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JULIOS,ANTHONY,J
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: BROWN,R
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 12/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A2
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: SENQUIZ

DATE/TIME OF TEST: 01/06/2018 20:01:56

Diagnostics Check		OK	19:55
Air Blank		0.000	19:55
Control Test		0.083	19:56
Air Blank		0.000	19:56
Subject Sample	# 1	0.000	19:57
Breath Volume		1.714	19:57
Air Blank		0.000	19:58
Air Blank		0.000	20:00
Subject Sample	# 2	0.000	20:00
Breath Volume		1.722	20:00
Air Blank		0.000	20:01
Control Test		0.080	20:01
Air Blank		0.000	20:01
Diagnostics Check		OK	20:01

Data Download Date/Time: 01/15/2018 16:18

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLLITO,BRAD,R
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MATHIS,JONATHAN
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 12/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A2
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 16:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:03

SUBJECT'S LAST NAME: LAYTON

DATE/TIME OF TEST: 01/07/2018 17:26:58

Diagnostics Check		OK	17:23	
Air Blank		0.000	17:23	
Control Test		0.082	17:23	
Air Blank		0.000	17:24	
Subject Sample	# 1		17:25	Subject Test Refused
Breath Volume		0.000	17:25	
Air Blank		0.000	17:26	
Control Test		0.082	17:26	
Air Blank		0.000	17:26	
Diagnostics Check		OK	17:26	

Data Download Date/Time: 01/15/2018 16:18

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STIMPSON,RACHEL,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: BROWN,RYAN
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 12/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A2
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: BURNS

DATE/TIME OF TEST: 01/08/2018 01:13:44

Diagnostics Check		OK	01:10	
Air Blank		0.000	01:11	
Control Test		0.081	01:11	
Air Blank		0.000	01:12	
Subject Sample	# 1		01:12	Subject Test Refused
Breath Volume		0.000	01:12	
Air Blank		0.000	01:12	
Control Test		0.082	01:13	
Air Blank		0.000	01:13	
Diagnostics Check		OK	01:13	

Data Download Date/Time: 01/15/2018 16:18

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COCHRANE,AARON,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: COCHRANE,AARON
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 12/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A2
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: MYERS

DATE/TIME OF TEST: 01/09/2018 19:39:25

Diagnostics Check		OK	19:30	
Air Blank		0.000	19:30	
Control Test		0.082	19:31	
Air Blank		0.000	19:31	
Subject Sample	# 1		19:34	No Sample Provided
Breath Volume		0.000	19:34	
Air Blank		0.000	19:35	
Air Blank		0.000	19:37	
Subject Sample	# 2		19:38	Subject Test Refused
Breath Volume		0.000	19:38	
Air Blank		0.000	19:38	
Control Test		0.081	19:38	
Air Blank		0.000	19:39	
Diagnostics Check		OK	19:39	

Data Download Date/Time: 01/15/2018 16:18

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COCHRANE,AARON,X
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: COCHRANE,AARON
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 12/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A2
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: FLEMING

DATE/TIME OF TEST: 01/13/2018 01:25:32

Diagnostics Check		OK	01:16
Air Blank		0.000	01:16
Control Test		0.080	01:16
Air Blank		0.000	01:17
Subject Sample	# 1	0.223	01:20
Breath Volume		2.191	01:20
Air Blank		0.000	01:21
Air Blank		0.000	01:23
Subject Sample	# 2	0.226	01:23
Breath Volume		2.593	01:23
Air Blank		0.000	01:24
Control Test		0.080	01:24
Air Blank		0.000	01:25
Diagnostics Check		OK	01:25

Data Download Date/Time: 01/15/2018 16:18

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO,I,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MORMANDO,D
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 12/23/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 17316080A2
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: KELLEY

DATE/TIME OF TEST: 01/14/2018 22:11:07

Diagnostics Check		OK	22:03
Air Blank		0.000	22:03
Control Test		0.082	22:03
Air Blank		0.000	22:04
Subject Sample	# 1	0.117	22:06
Breath Volume		2.316	22:06
Air Blank		0.000	22:07
Air Blank		0.000	22:08
Subject Sample	# 2	0.116	22:09
Breath Volume		1.871	22:09
Air Blank		0.000	22:10
Control Test		0.081	22:10
Air Blank		0.000	22:11
Diagnostics Check		OK	22:11

Data Download Date/Time: 01/15/2018 16:18

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COCHRANE,AARON,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: COCHRANE,AARON
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 12/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A2
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: BONILLA

DATE/TIME OF TEST: 01/15/2018 00:13:39

Diagnostics Check		OK	00:07
Air Blank		0.000	00:07
Control Test		0.083	00:08
Air Blank		0.000	00:08
Subject Sample	# 1	0.224	00:09
Breath Volume		1.117	00:09
Air Blank		0.000	00:09
Air Blank		0.000	00:11
Subject Sample	# 2	0.230	00:12
Breath Volume		1.148	00:12
Air Blank		0.000	00:12
Control Test		0.081	00:13
Air Blank		0.000	00:13
Diagnostics Check		OK	00:13

Data Download Date/Time: 01/15/2018 16:18

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,WILLIAM,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: PAYNE

DATE/TIME OF TEST: 01/01/2018 00:59:25

Diagnostics Check		OK	00:52
Air Blank		0.000	00:52
Control Test		0.082	00:52
Air Blank		0.000	00:53
Subject Sample	# 1	0.099	00:54
Breath Volume		2.109	00:54
Air Blank		0.000	00:54
Air Blank		0.000	00:56
Subject Sample	# 2	0.095	00:57
Breath Volume		1.253	00:57
Air Blank		0.000	00:58
Control Test		0.082	00:58
Air Blank		0.000	00:59
Diagnostics Check		OK	00:59

Data Download Date/Time: 01/25/2018 23:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,A,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: HOGAN

DATE/TIME OF TEST: 01/01/2018 03:37:36

Diagnostics Check		OK	03:26	
Air Blank		0.000	03:26	
Control Test		0.082	03:26	
Air Blank		0.000	03:27	
Subject Sample	# 1		03:30	No Sample Provided
Breath Volume		0.000	03:30	
Air Blank		0.000	03:31	
Air Blank		0.000	03:33	
Subject Sample	# 2		03:36	No Sample Provided
Breath Volume		0.000	03:36	
Air Blank		0.000	03:36	
Control Test		0.082	03:37	
Air Blank		0.000	03:37	
Diagnostics Check		OK	03:37	

Data Download Date/Time: 01/25/2018 23:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FRASER,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: LIPKIN

DATE/TIME OF TEST: 01/01/2018 05:05:43

Diagnostics Check		OK	04:59
Air Blank		0.000	04:59
Control Test		0.081	05:00
Air Blank		0.000	05:00
Subject Sample	# 1	0.163	05:01
Breath Volume		2.382	05:01
Air Blank		0.000	05:01
Air Blank		0.000	05:03
Subject Sample	# 2	0.166	05:04
Breath Volume		1.671	05:04
Air Blank		0.000	05:04
Control Test		0.081	05:05
Air Blank		0.000	05:05
Diagnostics Check		OK	05:05

Data Download Date/Time: 01/25/2018 23:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 04:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:17

SUBJECT'S LAST NAME: TERREY

DATE/TIME OF TEST: 01/01/2018 05:54:47

Diagnostics Check		OK	05:48
Air Blank		0.000	05:49
Control Test		0.081	05:49
Air Blank		0.000	05:50
Subject Sample	# 1	0.137	05:50
Breath Volume		2.425	05:50
Air Blank		0.000	05:50
Air Blank		0.000	05:52
Subject Sample	# 2	0.135	05:53
Breath Volume		2.382	05:53
Air Blank		0.000	05:53
Control Test		0.081	05:54
Air Blank		0.000	05:54
Diagnostics Check		OK	05:54

Data Download Date/Time: 01/25/2018 23:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,KYLE,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 21:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: PENDERGAST

DATE/TIME OF TEST: 01/02/2018 22:35:18

Diagnostics Check		OK	22:28
Air Blank		0.000	22:29
Control Test		0.083	22:29
Air Blank		0.000	22:29
Subject Sample	# 1	0.077	22:30
Breath Volume		1.300	22:30
Air Blank		0.000	22:31
Air Blank		0.000	22:33
Subject Sample	# 2	0.080	22:33
Breath Volume		1.835	22:33
Air Blank		0.000	22:34
Control Test		0.083	22:34
Air Blank		0.000	22:35
Diagnostics Check		OK	22:35

Data Download Date/Time: 01/25/2018 23:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,A,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,ALEX
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: STRYCZNY

DATE/TIME OF TEST: 01/04/2018 00:33:49

Diagnostics Check		OK	00:21	
Air Blank		0.000	00:22	
Control Test		0.083	00:22	
Air Blank		0.000	00:23	
Subject Sample	# 1		00:26	No Sample Provided
Breath Volume		0.000	00:26	
Air Blank		0.000	00:26	
Air Blank		0.000	00:28	
Subject Sample	# 2	0.173	00:31	Volume Not Met
Breath Volume		0.378	00:31	
Air Blank		0.000	00:32	
Control Test		0.083	00:33	
Air Blank		0.000	00:33	
Diagnostics Check		OK	00:33	

Data Download Date/Time: 01/25/2018 23:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEX,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: STRYCZNY

DATE/TIME OF TEST: 01/04/2018 00:44:17

Diagnostics Check		OK	00:34
Air Blank		0.000	00:35
Control Test		0.082	00:35
Air Blank		0.000	00:36
Subject Sample	# 1	0.176	00:38
Breath Volume		1.812	00:38
Air Blank		0.000	00:39
Air Blank		0.000	00:41
Subject Sample	# 2	0.173	00:42
Breath Volume		1.882	00:42
Air Blank		0.000	00:43
Control Test		0.080	00:43
Air Blank		0.000	00:44
Diagnostics Check		OK	00:44

Data Download Date/Time: 01/25/2018 23:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,ALEX
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: BALCH

DATE/TIME OF TEST: 01/04/2018 04:47:30

Diagnostics Check		OK	04:40
Air Blank		0.000	04:41
Control Test		0.084	04:41
Air Blank		0.000	04:42
Subject Sample	# 1	0.196	04:42
Breath Volume		1.785	04:42
Air Blank		0.000	04:43
Air Blank		0.000	04:44
Subject Sample	# 2	0.202	04:45
Breath Volume		2.097	04:45
Air Blank		0.000	04:46
Control Test		0.083	04:46
Air Blank		0.000	04:47
Diagnostics Check		OK	04:47

Data Download Date/Time: 01/25/2018 23:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBY,LARRY,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: JACOBY,LARRY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 21:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:43

SUBJECT'S LAST NAME: MANDY

DATE/TIME OF TEST: 01/04/2018 22:14:52

Diagnostics Check		OK	22:03	
Air Blank		0.000	22:03	
Control Test		0.084	22:04	
Air Blank		0.000	22:04	
Subject Sample	# 1		22:07	No Sample Provided
Breath Volume		0.000	22:07	
Air Blank		0.000	22:08	
Air Blank		0.000	22:10	
Subject Sample	# 2		22:13	No Sample Provided
Breath Volume		0.000	22:13	
Air Blank		0.000	22:13	
Control Test		0.083	22:14	
Air Blank		0.000	22:14	
Diagnostics Check		OK	22:14	

Data Download Date/Time: 01/25/2018 23:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORNING,PETER,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LORING,BLAINE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 14:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:23

SUBJECT'S LAST NAME: POLING

DATE/TIME OF TEST: 01/05/2018 16:11:07

Diagnostics Check		OK	16:04
Air Blank		0.000	16:05
Control Test		0.084	16:05
Air Blank		0.000	16:05
Subject Sample	# 1	0.000	16:06
Breath Volume		1.484	16:06
Air Blank		0.000	16:07
Air Blank		0.000	16:09
Subject Sample	# 2	0.000	16:09
Breath Volume		1.218	16:09
Air Blank		0.000	16:10
Control Test		0.083	16:10
Air Blank		0.000	16:11
Diagnostics Check		OK	16:11

Data Download Date/Time: 01/25/2018 23:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: FOSTER

DATE/TIME OF TEST: 01/06/2018 02:32:53

Diagnostics Check		OK	02:26
Air Blank		0.000	02:27
Control Test		0.083	02:27
Air Blank		0.000	02:28
Subject Sample	# 1	0.146	02:28
Breath Volume		2.062	02:28
Air Blank		0.000	02:29
Air Blank		0.000	02:30
Subject Sample	# 2	0.154	02:31
Breath Volume		1.636	02:31
Air Blank		0.000	02:31
Control Test		0.083	02:32
Air Blank		0.000	02:32
Diagnostics Check		OK	02:32

Data Download Date/Time: 01/25/2018 23:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONGMIRE,BENJAMIN,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HARTZIG,BOBBY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 04:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: MAYORGA RUANA

DATE/TIME OF TEST: 01/06/2018 05:57:30

Diagnostics Check		OK	05:51
Air Blank		0.000	05:51
Control Test		0.083	05:51
Air Blank		0.000	05:52
Subject Sample	# 1	0.184	05:53
Breath Volume		1.867	05:53
Air Blank		0.000	05:53
Air Blank		0.000	05:55
Subject Sample	# 2	0.197	05:55
Breath Volume		2.253	05:55
Air Blank		0.000	05:56
Control Test		0.082	05:56
Air Blank		0.000	05:57
Diagnostics Check		OK	05:57

Data Download Date/Time: 01/25/2018 23:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,AMANDA,C
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WARD,AMANDA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: MEZQUITE NICOLAS

DATE/TIME OF TEST: 01/07/2018 01:02:45

Diagnostics Check		OK	00:56
Air Blank		0.000	00:56
Control Test		0.083	00:57
Air Blank		0.000	00:57
Subject Sample	# 1	0.174	00:58
Breath Volume		1.285	00:58
Air Blank		0.000	00:58
Air Blank		0.000	01:00
Subject Sample	# 2	0.178	01:01
Breath Volume		1.890	01:01
Air Blank		0.000	01:01
Control Test		0.082	01:02
Air Blank		0.000	01:02
Diagnostics Check		OK	01:02

Data Download Date/Time: 01/25/2018 23:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORNING,PETER,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FIVECOAT,CURTIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 15:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:15

SUBJECT'S LAST NAME: LOPEZRUIZ

DATE/TIME OF TEST: 01/07/2018 16:51:33

Diagnostics Check		OK	16:39	
Air Blank		0.000	16:40	
Control Test		0.081	16:40	
Air Blank		0.000	16:41	
Subject Sample	# 1	0.192	16:44	Volume Not Met
Breath Volume		0.250	16:44	
Air Blank		0.000	16:44	
Air Blank		0.000	16:46	
Subject Sample	# 2	0.218	16:49	Volume Not Met
Breath Volume		0.250	16:49	
Air Blank		0.000	16:50	
Control Test		0.080	16:50	
Air Blank		0.000	16:51	
Diagnostics Check		OK	16:51	

Data Download Date/Time: 01/25/2018 23:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORNING,PETER,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FIVECOAT,CURTIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 15:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:15

SUBJECT'S LAST NAME: LOPEZRUIZ

DATE/TIME OF TEST: 01/07/2018 16:59:51

Diagnostics Check		OK	16:57	
Air Blank		0.000	16:57	
Control Test		0.080	16:57	
Air Blank		0.000	16:58	
Subject Sample	# 1		16:58	Subject Test Refused
Breath Volume		0.000	16:58	
Air Blank		0.000	16:58	
Control Test		0.081	16:59	
Air Blank		0.000	16:59	
Diagnostics Check		OK	16:59	

Data Download Date/Time: 01/25/2018 23:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,AMANDA,C
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WARD,AMANDA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: KANESKI

DATE/TIME OF TEST: 01/07/2018 22:04:02

Diagnostics Check		OK	22:00	
Air Blank		0.000	22:01	
Control Test		0.082	22:01	
Air Blank		0.000	22:01	
Subject Sample	# 1		22:02	RFI Detect
Breath Volume		0.000	22:02	
Air Blank		0.000	22:03	
Control Test		0.082	22:03	
Air Blank		0.000	22:03	
Diagnostics Check		OK	22:04	

Data Download Date/Time: 01/25/2018 23:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,AMANDA,C
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WARD,AMANDA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: KANESKI

DATE/TIME OF TEST: 01/07/2018 22:09:13

Diagnostics Check		OK	22:06	
Air Blank		0.000	22:06	
Control Test		0.083	22:06	
Air Blank		0.000	22:07	
Subject Sample	# 1		22:07	Subject Test Refused
Breath Volume		0.000	22:07	
Air Blank		0.000	22:08	
Control Test		0.082	22:08	
Air Blank		0.000	22:09	
Diagnostics Check		OK	22:09	

Data Download Date/Time: 01/25/2018 23:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,AMANDA,C
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MACKAIL,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 21:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: AMSTER

DATE/TIME OF TEST: 01/07/2018 22:23:06

Diagnostics Check		OK	22:20	
Air Blank		0.000	22:20	
Control Test		0.082	22:20	
Air Blank		0.000	22:21	
Subject Sample	# 1		22:21	Subject Test Refused
Breath Volume		0.000	22:21	
Air Blank		0.000	22:22	
Control Test		0.082	22:22	
Air Blank		0.000	22:23	
Diagnostics Check		OK	22:23	

Data Download Date/Time: 01/25/2018 23:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORR,RANDALL,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MITCHELL,ROBERT
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: INGRAM

DATE/TIME OF TEST: 01/08/2018 01:04:06

Diagnostics Check		OK	01:01	
Air Blank		0.000	01:01	
Control Test		0.083	01:01	
Air Blank		0.000	01:02	
Subject Sample	# 1		01:02	Subject Test Refused
Breath Volume		0.000	01:02	
Air Blank		0.000	01:03	
Control Test		0.083	01:03	
Air Blank		0.000	01:04	
Diagnostics Check		OK	01:04	

Data Download Date/Time: 01/25/2018 23:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBY,LARRY,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TAVERNIER,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: CHATMAN

DATE/TIME OF TEST: 01/09/2018 00:01:39

Diagnostics Check		OK	23:50	
Air Blank		0.000	23:50	
Control Test		0.082	23:51	
Air Blank		0.000	23:51	
Subject Sample	# 2		23:54	No Sample Provided
Breath Volume		0.000	23:54	
Air Blank		0.000	23:55	
Air Blank		0.000	23:57	
Subject Sample	# 1		00:00	No Sample Provided
Breath Volume		0.000	00:00	
Air Blank		0.000	00:00	
Control Test		0.082	00:01	
Air Blank		0.000	00:01	
Diagnostics Check		OK	00:01	

Data Download Date/Time: 01/25/2018 23:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTHONY,MICHAEL,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HANNA,JANE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 18:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:03

SUBJECT'S LAST NAME: GILLESPIE

DATE/TIME OF TEST: 01/11/2018 19:33:39

Diagnostics Check		OK	19:30	
Air Blank		0.000	19:31	
Control Test		0.082	19:31	
Air Blank		0.000	19:31	
Subject Sample	# 1		19:32	Subject Test Refused
Breath Volume		0.000	19:32	
Air Blank		0.000	19:32	
Control Test		0.082	19:33	
Air Blank		0.000	19:33	
Diagnostics Check		OK	19:33	

Data Download Date/Time: 01/25/2018 23:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: LANGSCHWAGER

DATE/TIME OF TEST: 01/12/2018 00:48:39

Diagnostics Check		OK	00:40
Air Blank		0.000	00:40
Control Test		0.082	00:40
Air Blank		0.000	00:41
Subject Sample	# 1	0.180	00:43
Breath Volume		1.363	00:43
Air Blank		0.000	00:44
Air Blank		0.000	00:46
Subject Sample	# 2	0.179	00:47
Breath Volume		1.601	00:47
Air Blank		0.000	00:47
Control Test		0.080	00:48
Air Blank		0.000	00:48
Diagnostics Check		OK	00:48

Data Download Date/Time: 01/25/2018 23:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REFUSE,RONY,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: REFUSE,RONY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: M450770620420

DATE/TIME OF TEST: 01/12/2018 03:55:31

Diagnostics Check		OK	03:40	
Air Blank		0.000	03:41	
Control Test		0.082	03:41	
Air Blank		0.000	03:42	
Subject Sample	# 1	0.314	03:45	
Breath Volume		1.792	03:45	
Air Blank		0.000	03:45	
Air Blank		0.000	03:47	
Subject Sample	# 2	0.275	03:50	Volume Not Met
Breath Volume		0.269	03:50	
Air Blank		0.000	03:51	
Air Blank		0.000	03:53	
Subject Sample	# 3	0.308	03:53	
Breath Volume		1.371	03:53	
Air Blank		0.000	03:54	
Control Test		0.077	03:54	
Air Blank		0.000	03:55	
Diagnostics Check		OK	03:55	

Data Download Date/Time: 01/25/2018 23:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: ARRUDA

DATE/TIME OF TEST: 01/13/2018 03:33:07

Diagnostics Check		OK	03:30	
Air Blank		0.000	03:30	
Control Test		0.082	03:30	
Air Blank		0.000	03:31	
Subject Sample	# 1		03:31	Subject Test Refused
Breath Volume		0.000	03:31	
Air Blank		0.000	03:32	
Control Test		0.082	03:32	
Air Blank		0.000	03:33	
Diagnostics Check		OK	03:33	

Data Download Date/Time: 01/25/2018 23:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: LIBBY

DATE/TIME OF TEST: 01/14/2018 03:49:40

Diagnostics Check		OK	03:43
Air Blank		0.000	03:44
Control Test		0.083	03:44
Air Blank		0.000	03:44
Subject Sample	# 1	0.196	03:45
Breath Volume		1.515	03:45
Air Blank		0.000	03:45
Air Blank		0.000	03:47
Subject Sample	# 2	0.191	03:48
Breath Volume		1.917	03:48
Air Blank		0.000	03:48
Control Test		0.082	03:49
Air Blank		0.000	03:49
Diagnostics Check		OK	03:49

Data Download Date/Time: 01/25/2018 23:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORNING,PETER,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CAREY,KEVIN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 11:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:59

SUBJECT'S LAST NAME: TRZASKA

DATE/TIME OF TEST: 01/15/2018 12:27:53

Diagnostics Check		OK	12:21
Air Blank		0.000	12:22
Control Test		0.083	12:22
Air Blank		0.000	12:22
Subject Sample	# 1	0.000	12:23
Breath Volume		1.445	12:23
Air Blank		0.000	12:23
Air Blank		0.000	12:25
Subject Sample	# 2	0.000	12:26
Breath Volume		1.312	12:26
Air Blank		0.000	12:26
Control Test		0.083	12:27
Air Blank		0.000	12:27
Diagnostics Check		OK	12:27

Data Download Date/Time: 01/25/2018 23:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DONOHUE,LAWRENCE,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DONOHUE,LAWRENCE
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 04:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:22

SUBJECT'S LAST NAME: WESTFALL

DATE/TIME OF TEST: 01/16/2018 17:49:59

Diagnostics Check		OK	17:42
Air Blank		0.000	17:42
Control Test		0.082	17:43
Air Blank		0.000	17:43
Subject Sample	# 1	0.245	17:45
Breath Volume		1.769	17:45
Air Blank		0.000	17:45
Air Blank		0.000	17:47
Subject Sample	# 2	0.247	17:48
Breath Volume		2.472	17:48
Air Blank		0.000	17:49
Control Test		0.081	17:49
Air Blank		0.000	17:49
Diagnostics Check		OK	17:49

Data Download Date/Time: 01/25/2018 23:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REFUSE,RONY,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: REFUSE,RONY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 18:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:28

SUBJECT'S LAST NAME: GUERRERO

DATE/TIME OF TEST: 01/18/2018 19:59:26

Diagnostics Check		OK	19:48
Air Blank		0.000	19:49
Control Test		0.083	19:49
Air Blank		0.000	19:49
Subject Sample	# 1	0.142	19:52
Breath Volume		1.144	19:52
Air Blank		0.000	19:52
Air Blank		0.000	19:54
Subject Sample	# 2	0.135	19:57
Breath Volume		1.257	19:57
Air Blank		0.000	19:58
Control Test		0.083	19:58
Air Blank		0.000	19:59
Diagnostics Check		OK	19:59

Data Download Date/Time: 01/25/2018 23:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTHONY,MICHAEL,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SKALKO,RYAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: WITT

DATE/TIME OF TEST: 01/18/2018 23:12:46

Diagnostics Check		OK	23:05
Air Blank		0.000	23:05
Control Test		0.080	23:05
Air Blank		0.000	23:06
Subject Sample	# 1	0.214	23:07
Breath Volume		2.343	23:07
Air Blank		0.000	23:07
Air Blank		0.000	23:09
Subject Sample	# 2	0.208	23:11
Breath Volume		1.531	23:11
Air Blank		0.000	23:11
Control Test		0.083	23:12
Air Blank		0.000	23:12
Diagnostics Check		OK	23:12

Data Download Date/Time: 01/25/2018 23:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,WILLIAM,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 01/20/2018 23:07:17

Diagnostics Check		OK	22:55	
Air Blank		0.000	22:56	
Control Test		0.083	22:56	
Air Blank		0.000	22:57	
Subject Sample	# 1		23:00	No Sample Provided
Breath Volume		0.000	23:00	
Air Blank		0.000	23:00	
Air Blank		0.000	23:02	
Subject Sample	# 2		23:05	No Sample Provided
Breath Volume		0.000	23:05	
Air Blank		0.000	23:06	
Control Test		0.083	23:06	
Air Blank		0.000	23:07	
Diagnostics Check		OK	23:07	

Data Download Date/Time: 01/25/2018 23:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTHONY,MICHAEL,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: POLLOCK

DATE/TIME OF TEST: 01/21/2018 01:14:48

Diagnostics Check		OK	01:11	
Air Blank		0.000	01:12	
Control Test		0.083	01:12	
Air Blank		0.000	01:13	
Subject Sample	# 1		01:13	Subject Test Refused
Breath Volume		0.000	01:13	
Air Blank		0.000	01:13	
Control Test		0.081	01:14	
Air Blank		0.000	01:14	
Diagnostics Check		OK	01:14	

Data Download Date/Time: 01/25/2018 23:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTHONY,MICHAEL,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: NESTOR,ERIN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: ELLISON

DATE/TIME OF TEST: 01/21/2018 02:05:03

Diagnostics Check		OK	02:02	
Air Blank		0.000	02:02	
Control Test		0.083	02:02	
Air Blank		0.000	02:03	
Subject Sample	# 1		02:03	Subject Test Refused
Breath Volume		0.000	02:03	
Air Blank		0.000	02:04	
Control Test		0.083	02:04	
Air Blank		0.000	02:04	
Diagnostics Check		OK	02:05	

Data Download Date/Time: 01/25/2018 23:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BICKFORD,BRANT,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: STILES,SHANE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:57

SUBJECT'S LAST NAME: BUGNA

DATE/TIME OF TEST: 01/21/2018 19:26:49

Diagnostics Check		OK	19:18
Air Blank		0.000	19:19
Control Test		0.082	19:19
Air Blank		0.000	19:20
Subject Sample	# 1	0.261	19:22
Breath Volume		1.539	19:22
Air Blank		0.000	19:23
Air Blank		0.000	19:24
Subject Sample	# 2	0.251	19:25
Breath Volume		1.796	19:25
Air Blank		0.000	19:25
Control Test		0.080	19:26
Air Blank		0.000	19:26
Diagnostics Check		OK	19:26

Data Download Date/Time: 01/25/2018 23:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REFUSE,RONY,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: REFUSE,RONY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 19:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:17

SUBJECT'S LAST NAME: HILLEY

DATE/TIME OF TEST: 01/21/2018 20:46:54

Diagnostics Check		OK	20:39
Air Blank		0.000	20:39
Control Test		0.082	20:39
Air Blank		0.000	20:40
Subject Sample	# 1	0.136	20:41
Breath Volume		1.109	20:41
Air Blank		0.000	20:42
Air Blank		0.000	20:44
Subject Sample	# 2	0.149	20:45
Breath Volume		1.785	20:45
Air Blank		0.000	20:45
Control Test		0.081	20:46
Air Blank		0.000	20:46
Diagnostics Check		OK	20:46

Data Download Date/Time: 01/25/2018 23:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BICKFORD,BRANT,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 22:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: JACOBS

DATE/TIME OF TEST: 01/21/2018 23:00:04

Diagnostics Check		OK	22:57	
Air Blank		0.000	22:57	
Control Test		0.081	22:57	
Air Blank		0.000	22:58	
Subject Sample	# 1		22:58	Subject Test Refused
Breath Volume		0.000	22:58	
Air Blank		0.000	22:59	
Control Test		0.082	22:59	
Air Blank		0.000	22:59	
Diagnostics Check		OK	23:00	

Data Download Date/Time: 01/25/2018 23:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BICKFORD,BRANT,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LUDTKE,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: GILMORE

DATE/TIME OF TEST: 01/21/2018 23:32:11

Diagnostics Check		OK	23:29	
Air Blank		0.000	23:29	
Control Test		0.081	23:30	
Air Blank		0.000	23:30	
Subject Sample	# 1		23:30	Subject Test Refused
Breath Volume		0.000	23:30	
Air Blank		0.000	23:31	
Control Test		0.082	23:31	
Air Blank		0.000	23:32	
Diagnostics Check		OK	23:32	

Data Download Date/Time: 01/25/2018 23:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIBLEY,TRAVIS,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SIBLEY,TRAVIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 14:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:16

SUBJECT'S LAST NAME: SHRIVER

DATE/TIME OF TEST: 01/23/2018 15:42:14

Diagnostics Check		OK	15:36
Air Blank		0.000	15:36
Control Test		0.082	15:37
Air Blank		0.000	15:37
Subject Sample	# 1	0.000	15:38
Breath Volume		1.195	15:38
Air Blank		0.000	15:38
Air Blank		0.000	15:40
Subject Sample	# 2	0.000	15:40
Breath Volume		1.468	15:40
Air Blank		0.000	15:41
Control Test		0.082	15:41
Air Blank		0.000	15:42
Diagnostics Check		OK	15:42

Data Download Date/Time: 01/25/2018 23:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DONOHUE,S
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: HOOPINGARNER

DATE/TIME OF TEST: 01/25/2018 00:28:16

Diagnostics Check		OK	00:25	
Air Blank		0.000	00:25	
Control Test		0.083	00:26	
Air Blank		0.000	00:26	
Subject Sample	# 1		00:26	Subject Test Refused
Breath Volume		0.000	00:26	
Air Blank		0.000	00:27	
Control Test		0.083	00:27	
Air Blank		0.000	00:28	
Diagnostics Check		OK	00:28	

Data Download Date/Time: 01/25/2018 23:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: BUCHANAN

DATE/TIME OF TEST: 01/25/2018 03:35:56

Diagnostics Check		OK	03:27	
Air Blank		0.000	03:27	
Control Test		0.083	03:28	
Air Blank		0.000	03:28	
Subject Sample	# 1		03:31	No Sample Provided
Breath Volume		0.000	03:31	
Air Blank		0.000	03:32	
Air Blank		0.000	03:34	
Subject Sample	# 2		03:34	Subject Test Refused
Breath Volume		0.000	03:34	
Air Blank		0.000	03:35	
Control Test		0.083	03:35	
Air Blank		0.000	03:35	
Diagnostics Check		OK	03:35	

Data Download Date/Time: 01/25/2018 23:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: CHARLES

DATE/TIME OF TEST: 01/26/2018 02:10:22

Diagnostics Check		OK	01:58	
Air Blank		0.000	01:59	
Control Test		0.082	01:59	
Air Blank		0.000	02:00	
Subject Sample	# 1		02:03	No Sample Provided
Breath Volume		0.000	02:03	
Air Blank		0.000	02:03	
Air Blank		0.000	02:05	
Subject Sample	# 2		02:09	No Sample Provided
Breath Volume		0.000	02:09	
Air Blank		0.000	02:09	
Control Test		0.083	02:09	
Air Blank		0.000	02:10	
Diagnostics Check		OK	02:10	

Data Download Date/Time: 02/16/2018 23:56

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,KYLE,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GRIFFIN,KYLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: LOPEZ-RIVERA

DATE/TIME OF TEST: 01/26/2018 03:56:44

Diagnostics Check		OK	03:46	
Air Blank		0.000	03:46	
Control Test		0.082	03:47	
Air Blank		0.000	03:47	
Subject Sample	# 1		03:50	No Sample Provided
Breath Volume		0.000	03:50	
Air Blank		0.000	03:51	
Air Blank		0.000	03:53	
Subject Sample	# 2		03:55	Subject Test Refused
Breath Volume		0.000	03:55	
Air Blank		0.000	03:55	
Control Test		0.082	03:56	
Air Blank		0.000	03:56	
Diagnostics Check		OK	03:56	

Data Download Date/Time: 02/16/2018 23:56

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 21:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 01/26/2018 22:16:32

Diagnostics Check		OK	22:09
Air Blank		0.000	22:10
Control Test		0.082	22:10
Air Blank		0.000	22:11
Subject Sample	# 1	0.149	22:11
Breath Volume		3.449	22:11
Air Blank		0.000	22:12
Air Blank		0.000	22:14
Subject Sample	# 2	0.144	22:15
Breath Volume		3.472	22:15
Air Blank		0.000	22:15
Control Test		0.080	22:15
Air Blank		0.000	22:16
Diagnostics Check		OK	22:16

Data Download Date/Time: 02/16/2018 23:56

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GEOGHEGAN,L
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 01/27/2018 00:09:51

Diagnostics Check		OK	00:01
Air Blank		0.000	00:02
Control Test		0.083	00:02
Air Blank		0.000	00:03
Subject Sample	# 1	0.226	00:04
Breath Volume		3.320	00:04
Air Blank		0.000	00:05
Air Blank		0.000	00:06
Subject Sample	# 2	0.232	00:08
Breath Volume		1.558	00:08
Air Blank		0.000	00:08
Control Test		0.081	00:09
Air Blank		0.000	00:09
Diagnostics Check		OK	00:09

Data Download Date/Time: 02/16/2018 23:56

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: PRESBURY

DATE/TIME OF TEST: 01/27/2018 02:51:19

Diagnostics Check		OK	02:44
Air Blank		0.000	02:44
Control Test		0.082	02:44
Air Blank		0.000	02:45
Subject Sample	# 1	0.115	02:46
Breath Volume		2.460	02:46
Air Blank		0.000	02:46
Air Blank		0.000	02:48
Subject Sample	# 2	0.123	02:49
Breath Volume		2.500	02:49
Air Blank		0.000	02:50
Control Test		0.083	02:50
Air Blank		0.000	02:51
Diagnostics Check		OK	02:51

Data Download Date/Time: 02/16/2018 23:56

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: HOWELL

DATE/TIME OF TEST: 01/27/2018 03:36:30

Diagnostics Check		OK	03:30
Air Blank		0.000	03:30
Control Test		0.082	03:30
Air Blank		0.000	03:31
Subject Sample	# 1	0.185	03:31
Breath Volume		1.609	03:31
Air Blank		0.000	03:32
Air Blank		0.000	03:34
Subject Sample	# 2	0.196	03:34
Breath Volume		1.464	03:34
Air Blank		0.000	03:35
Control Test		0.082	03:35
Air Blank		0.000	03:36
Diagnostics Check		OK	03:36

Data Download Date/Time: 02/16/2018 23:56

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBY,LARRY,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: JACOBY,LARRY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 19:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:13

SUBJECT'S LAST NAME: WEISE

DATE/TIME OF TEST: 01/27/2018 20:40:59

Diagnostics Check		OK	20:34
Air Blank		0.000	20:35
Control Test		0.082	20:35
Air Blank		0.000	20:35
Subject Sample	# 1	0.182	20:36
Breath Volume		2.101	20:36
Air Blank		0.000	20:37
Air Blank		0.000	20:38
Subject Sample	# 2	0.191	20:39
Breath Volume		2.160	20:39
Air Blank		0.000	20:40
Control Test		0.082	20:40
Air Blank		0.000	20:40
Diagnostics Check		OK	20:40

Data Download Date/Time: 02/16/2018 23:56

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAIR,RONALD,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAIR,RONALD
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: BILLINGTON

DATE/TIME OF TEST: 01/28/2018 00:39:40

Diagnostics Check		OK	00:36	
Air Blank		0.000	00:36	
Control Test		0.082	00:37	
Air Blank		0.000	00:37	
Subject Sample	# 1		00:38	Subject Test Refused
Breath Volume		0.000	00:38	
Air Blank		0.000	00:38	
Control Test		0.082	00:39	
Air Blank		0.000	00:39	
Diagnostics Check		OK	00:39	

Data Download Date/Time: 02/16/2018 23:56

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FRASER,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: LANDRY

DATE/TIME OF TEST: 01/28/2018 02:23:25

Diagnostics Check		OK	02:16
Air Blank		0.000	02:16
Control Test		0.082	02:17
Air Blank		0.000	02:17
Subject Sample	# 1	0.152	02:19
Breath Volume		1.492	02:19
Air Blank		0.000	02:19
Air Blank		0.000	02:21
Subject Sample	# 2	0.149	02:21
Breath Volume		1.671	02:21
Air Blank		0.000	02:22
Control Test		0.082	02:22
Air Blank		0.000	02:23
Diagnostics Check		OK	02:23

Data Download Date/Time: 02/16/2018 23:56

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ROSADO,I
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: HOGAN

DATE/TIME OF TEST: 01/28/2018 03:21:09

Diagnostics Check	OK	03:20	
Air Blank		03:20	RFI Detect
Air Blank	0.000	03:21	

Data Download Date/Time: 02/16/2018 23:56

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ROSADO,I
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: HOGAN

DATE/TIME OF TEST: 01/28/2018 03:29:27

Diagnostics Check		OK	03:22
Air Blank		0.000	03:23
Control Test		0.082	03:23
Air Blank		0.000	03:23
Subject Sample	# 1	0.179	03:24
Breath Volume		1.199	03:24
Air Blank		0.000	03:25
Air Blank		0.000	03:27
Subject Sample	# 2	0.189	03:27
Breath Volume		1.156	03:27
Air Blank		0.000	03:28
Control Test		0.082	03:28
Air Blank		0.000	03:29
Diagnostics Check		OK	03:29

Data Download Date/Time: 02/16/2018 23:56

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 01/28/2018 05:05:21

Diagnostics Check		OK	04:56	
Air Blank		0.000	04:56	
Control Test		0.082	04:57	
Air Blank		0.000	04:57	
Subject Sample	# 1		05:00	No Sample Provided
Breath Volume		0.000	05:00	
Air Blank		0.000	05:01	
Air Blank		0.000	05:03	
Subject Sample	# 2		05:03	Subject Test Refused
Breath Volume		0.000	05:03	
Air Blank		0.000	05:04	
Control Test		0.082	05:04	
Air Blank		0.000	05:05	
Diagnostics Check		OK	05:05	

Data Download Date/Time: 02/16/2018 23:56

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,KYLE,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GRIFFIN,KYLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 04:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:08

SUBJECT'S LAST NAME: JAMES

DATE/TIME OF TEST: 01/30/2018 05:31:56

Diagnostics Check		OK	05:28	
Air Blank		0.000	05:28	
Control Test		0.083	05:29	
Air Blank		0.000	05:29	
Subject Sample	# 1		05:30	Subject Test Refused
Breath Volume		0.000	05:30	
Air Blank		0.000	05:31	
Control Test		0.083	05:31	
Air Blank		0.000	05:31	
Diagnostics Check		OK	05:31	

Data Download Date/Time: 02/16/2018 23:56

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: MENDES

DATE/TIME OF TEST: 01/31/2018 23:20:26

Diagnostics Check		OK	23:14
Air Blank		0.000	23:14
Control Test		0.083	23:15
Air Blank		0.000	23:15
Subject Sample	# 1	0.022	23:16
Breath Volume		3.136	23:16
Air Blank		0.000	23:16
Air Blank		0.000	23:18
Subject Sample	# 2	0.022	23:18
Breath Volume		3.367	23:18
Air Blank		0.000	23:19
Control Test		0.083	23:19
Air Blank		0.000	23:20
Diagnostics Check		OK	23:20

Data Download Date/Time: 02/16/2018 23:56

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORRIS,DWIGHT,
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: NORRIS,DWIGHT
ARRESTING OFFICER AGENCY: Perry PD
LAST AGENCY INSPECTION: 12/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: COURT ORD
ARREST TIME: 11:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:00

SUBJECT'S LAST NAME: MURRAY

DATE/TIME OF TEST: 01/04/2018 11:35:34

Diagnostics Check		OK	11:29
Air Blank		0.000	11:29
Control Test		0.083	11:30
Air Blank		0.000	11:30
Subject Sample	# 1	0.000	11:31
Breath Volume		2.179	11:31
Air Blank		0.000	11:31
Air Blank		0.000	11:33
Subject Sample	# 2	0.000	11:34
Breath Volume		3.128	11:34
Air Blank		0.000	11:34
Control Test		0.081	11:34
Air Blank		0.000	11:35
Diagnostics Check		OK	11:35

Data Download Date/Time: 01/29/2018 15:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARNOLD,JOHN,S
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: THIGPEN,CHRISTOPHER
ARRESTING OFFICER AGENCY: Taylor County SO
LAST AGENCY INSPECTION: 12/08/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: DUI
ARREST TIME: 18:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 01/05/2018 19:11:37

Diagnostics Check		OK	19:05
Air Blank		0.000	19:05
Control Test		0.083	19:05
Air Blank		0.000	19:06
Subject Sample	# 1	0.257	19:07
Breath Volume		3.273	19:07
Air Blank		0.000	19:07
Air Blank		0.000	19:09
Subject Sample	# 2	0.264	19:10
Breath Volume		3.164	19:10
Air Blank		0.000	19:10
Control Test		0.082	19:11
Air Blank		0.000	19:11
Diagnostics Check		OK	19:11

Data Download Date/Time: 01/29/2018 15:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWINDLE,CHARLES,M
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: SWINDLE,CHARLES
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 12/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: HEAD

DATE/TIME OF TEST: 01/08/2018 22:39:12

Diagnostics Check		OK	22:33
Air Blank		0.000	22:33
Control Test		0.081	22:33
Air Blank		0.000	22:34
Subject Sample	# 1	0.082	22:34
Breath Volume		2.398	22:34
Air Blank		0.000	22:35
Air Blank		0.000	22:37
Subject Sample	# 2	0.080	22:37
Breath Volume		2.335	22:37
Air Blank		0.000	22:38
Control Test		0.081	22:38
Air Blank		0.000	22:39
Diagnostics Check		OK	22:39

Data Download Date/Time: 01/29/2018 15:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICKETSON,JAY,R
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: RICKETSON,J
ARRESTING OFFICER AGENCY: Perry PD
LAST AGENCY INSPECTION: 12/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: PROB/PAROL
ARREST TIME: 10:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:00

SUBJECT'S LAST NAME: MURRAY

DATE/TIME OF TEST: 01/10/2018 10:35:01

Diagnostics Check		OK	10:28
Air Blank		0.000	10:29
Control Test		0.081	10:29
Air Blank		0.000	10:30
Subject Sample	# 1	0.000	10:30
Breath Volume		1.253	10:30
Air Blank		0.000	10:31
Air Blank		0.000	10:33
Subject Sample	# 2	0.000	10:33
Breath Volume		3.101	10:33
Air Blank		0.000	10:34
Control Test		0.081	10:34
Air Blank		0.000	10:34
Diagnostics Check		OK	10:35

Data Download Date/Time: 01/29/2018 15:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARNOLD,JOHN,S
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: MURPHY,TIMMY
ARRESTING OFFICER AGENCY: Perry PD
LAST AGENCY INSPECTION: 12/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 01/16/2018 01:59:19

Diagnostics Check		OK	01:43	
Air Blank		0.000	01:43	
Control Test		0.083	01:44	
Air Blank		0.000	01:44	
Subject Sample	# 1	0.196	01:46	
Breath Volume		1.710	01:46	
Air Blank		0.000	01:47	
Air Blank		0.000	01:48	
Subject Sample	# 2	0.175	01:52	Volume Not Met
Breath Volume		0.000	01:52	
Air Blank		0.000	01:52	
Air Blank		0.000	01:54	
Subject Sample	# 3	0.175	01:57	Volume Not Met
Breath Volume		0.480	01:57	
Air Blank		0.000	01:58	
Control Test		0.080	01:58	
Air Blank		0.000	01:59	
Diagnostics Check		OK	01:59	

Data Download Date/Time: 01/29/2018 15:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CEPHUS,JAMIN,I
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: CEPHUS,JAMIN
ARRESTING OFFICER AGENCY: Perry PD
LAST AGENCY INSPECTION: 12/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: COURT ORD
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:30

SUBJECT'S LAST NAME: MURRAY

DATE/TIME OF TEST: 01/18/2018 10:59:35

Diagnostics Check		OK	10:53
Air Blank		0.000	10:53
Control Test		0.082	10:54
Air Blank		0.000	10:54
Subject Sample	# 1	0.000	10:55
Breath Volume		3.890	10:55
Air Blank		0.000	10:55
Air Blank		0.000	10:57
Subject Sample	# 2	0.000	10:58
Breath Volume		3.609	10:58
Air Blank		0.000	10:58
Control Test		0.082	10:58
Air Blank		0.000	10:59
Diagnostics Check		OK	10:59

Data Download Date/Time: 01/29/2018 15:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICKETSON,JAY,R
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: RICKETSON,J
ARRESTING OFFICER AGENCY: Perry PD
LAST AGENCY INSPECTION: 12/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: ADMIN
ARREST TIME: 09:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:10

SUBJECT'S LAST NAME: DOUGHTERY

DATE/TIME OF TEST: 01/19/2018 09:44:13

Diagnostics Check		OK	09:37
Air Blank		0.000	09:38
Control Test		0.082	09:38
Air Blank		0.000	09:39
Subject Sample	# 1	0.000	09:39
Breath Volume		3.726	09:39
Air Blank		0.000	09:40
Air Blank		0.000	09:42
Subject Sample	# 2	0.000	09:42
Breath Volume		3.925	09:42
Air Blank		0.000	09:43
Control Test		0.082	09:43
Air Blank		0.000	09:44
Diagnostics Check		OK	09:44

Data Download Date/Time: 01/29/2018 15:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICKETSON,JAY,R
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: RICKETSON,J
ARRESTING OFFICER AGENCY: Perry PD
LAST AGENCY INSPECTION: 12/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: PROB/PAROL
ARREST TIME: 10:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:20

SUBJECT'S LAST NAME: M600436942560

DATE/TIME OF TEST: 01/26/2018 10:53:08

Diagnostics Check		OK	10:47
Air Blank		0.000	10:47
Control Test		0.082	10:47
Air Blank		0.000	10:48
Subject Sample	# 1	0.000	10:48
Breath Volume		2.863	10:48
Air Blank		0.000	10:49
Air Blank		0.000	10:51
Subject Sample	# 2	0.000	10:51
Breath Volume		2.292	10:51
Air Blank		0.000	10:52
Control Test		0.081	10:52
Air Blank		0.000	10:53
Diagnostics Check		OK	10:53

Data Download Date/Time: 01/29/2018 15:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CEPHUS,JAMIN,I
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: CEPHUS,JAMIN
ARRESTING OFFICER AGENCY: Perry PD
LAST AGENCY INSPECTION: 12/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: SAYLOR

DATE/TIME OF TEST: 01/26/2018 23:42:50

Diagnostics Check	OK	23:41	
Air Blank	0.000	23:41	
Control Test	0.081	23:42	
Air Blank		23:42	RFI Detect
Air Blank	0.000	23:42	

Data Download Date/Time: 01/29/2018 15:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CEPHUS,JAMIN,I
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: CEPHUS,JAMIN
ARRESTING OFFICER AGENCY: Perry PD
LAST AGENCY INSPECTION: 12/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: SAYLOR

DATE/TIME OF TEST: 01/26/2018 23:51:19

Diagnostics Check		OK	23:44
Air Blank		0.000	23:45
Control Test		0.081	23:45
Air Blank		0.000	23:46
Subject Sample	# 1	0.000	23:46
Breath Volume		4.734	23:46
Air Blank		0.000	23:47
Air Blank		0.000	23:49
Subject Sample	# 2	0.000	23:49
Breath Volume		3.332	23:49
Air Blank		0.000	23:50
Control Test		0.081	23:50
Air Blank		0.000	23:51
Diagnostics Check		OK	23:51

Data Download Date/Time: 01/29/2018 15:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CEPHUS,JAMIN,I
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: CEPHUS,JAMIN
ARRESTING OFFICER AGENCY: Perry PD
LAST AGENCY INSPECTION: 12/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 01/27/2018 23:18:40

Diagnostics Check		OK	23:12
Air Blank		0.000	23:12
Control Test		0.081	23:12
Air Blank		0.000	23:13
Subject Sample	# 1	0.189	23:14
Breath Volume		3.328	23:14
Air Blank		0.000	23:14
Air Blank		0.000	23:16
Subject Sample	# 2	0.186	23:17
Breath Volume		3.457	23:17
Air Blank		0.000	23:17
Control Test		0.082	23:18
Air Blank		0.000	23:18
Diagnostics Check		OK	23:18

Data Download Date/Time: 01/29/2018 15:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARNOLD,JOHN,S
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: ARNOLD,JOHN
ARRESTING OFFICER AGENCY: Perry PD
LAST AGENCY INSPECTION: 01/29/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: ADMIN
ARREST TIME: 10:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:35

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 01/30/2018 11:04:26

Diagnostics Check		OK	10:57
Air Blank		0.000	10:58
Control Test		0.082	10:58
Air Blank		0.000	10:59
Subject Sample	# 1	0.000	10:59
Breath Volume		1.453	10:59
Air Blank		0.000	11:00
Air Blank		0.000	11:02
Subject Sample	# 2	0.000	11:03
Breath Volume		4.718	11:03
Air Blank		0.000	11:03
Control Test		0.082	11:03
Air Blank		0.000	11:04
Diagnostics Check		OK	11:04

Data Download Date/Time: 02/26/2018 15:23

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARNOLD,JOHN,S
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: ARNOLD,JOHN
ARRESTING OFFICER AGENCY: Perry PD
LAST AGENCY INSPECTION: 01/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: PROB/PAROL
ARREST TIME: 11:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:15

SUBJECT'S LAST NAME: MURRAY

DATE/TIME OF TEST: 01/30/2018 11:42:25

Diagnostics Check		OK	11:36
Air Blank		0.000	11:36
Control Test		0.082	11:37
Air Blank		0.000	11:37
Subject Sample	# 1	0.000	11:38
Breath Volume		3.023	11:38
Air Blank		0.000	11:38
Air Blank		0.000	11:40
Subject Sample	# 2	0.000	11:41
Breath Volume		2.539	11:41
Air Blank		0.000	11:41
Control Test		0.080	11:41
Air Blank		0.000	11:42
Diagnostics Check		OK	11:42

Data Download Date/Time: 02/26/2018 15:23

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLI,ANTHONY,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SMITH,EMILY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: SHAW

DATE/TIME OF TEST: 01/28/2018 04:11:00

Diagnostics Check		OK	03:59	
Air Blank		0.000	04:00	
Control Test		0.081	04:00	
Air Blank		0.000	04:00	
Subject Sample	# 1	0.000	04:04	Volume Not Met
Breath Volume		0.019	04:04	
Air Blank		0.000	04:04	
Air Blank		0.000	04:06	
Subject Sample	# 2		04:09	No Sample Provided
Breath Volume		0.000	04:09	
Air Blank		0.000	04:10	
Control Test		0.080	04:10	
Air Blank		0.000	04:10	
Diagnostics Check		OK	04:11	

Data Download Date/Time: 02/05/2018 10:39

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLI,ANTHONY,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SMITH,EMILY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: SHAW

DATE/TIME OF TEST: 01/28/2018 04:15:46

Diagnostics Check		OK	04:12	
Air Blank		0.000	04:13	
Control Test		0.081	04:13	
Air Blank		0.000	04:14	
Subject Sample	# 1		04:14	Subject Test Refused
Breath Volume		0.000	04:14	
Air Blank		0.000	04:14	
Control Test		0.081	04:15	
Air Blank		0.000	04:15	
Diagnostics Check		OK	04:15	

Data Download Date/Time: 02/05/2018 10:39

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005454

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOEDICKE,WILLIAM,R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: JOEDICKE,WILLIAM
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 12/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: MICKETT

DATE/TIME OF TEST: 01/10/2018 21:15:34

Diagnostics Check		OK	21:08
Air Blank		0.000	21:09
Control Test		0.076	21:09
Air Blank		0.000	21:10
Subject Sample	# 1	0.198	21:10
Breath Volume		2.820	21:10
Air Blank		0.000	21:11
Air Blank		0.000	21:13
Subject Sample	# 2	0.181	21:13
Breath Volume		1.218	21:13
Air Blank		0.000	21:14
Control Test		0.076	21:14
Air Blank		0.000	21:15
Diagnostics Check		OK	21:15

Data Download Date/Time: 01/31/2018 12:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005454

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOEDICKE,WILLIAM,R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: JOEDICKE,WILLIAM
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 12/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: BELLAMY

DATE/TIME OF TEST: 01/12/2018 23:12:32

Diagnostics Check		OK	23:05
Air Blank		0.000	23:06
Control Test		0.078	23:06
Air Blank		0.000	23:07
Subject Sample	# 1	0.134	23:07
Breath Volume		2.160	23:07
Air Blank		0.000	23:08
Air Blank		0.000	23:10
Subject Sample	# 2	0.126	23:10
Breath Volume		2.062	23:10
Air Blank		0.000	23:11
Control Test		0.077	23:11
Air Blank		0.000	23:12
Diagnostics Check		OK	23:12

Data Download Date/Time: 01/31/2018 12:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005454

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOEDICKE,WILLIAM,R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: JOEDICKE,WILLIAM
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 12/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: FICK

DATE/TIME OF TEST: 01/13/2018 03:26:15

Diagnostics Check		OK	03:19
Air Blank		0.000	03:20
Control Test		0.078	03:20
Air Blank		0.000	03:21
Subject Sample	# 1	0.087	03:21
Breath Volume		1.886	03:21
Air Blank		0.000	03:22
Air Blank		0.000	03:24
Subject Sample	# 2	0.087	03:24
Breath Volume		3.011	03:24
Air Blank		0.000	03:25
Control Test		0.077	03:25
Air Blank		0.000	03:26
Diagnostics Check		OK	03:26

Data Download Date/Time: 01/31/2018 12:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005454

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOEDICKE,WILLIAM,R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: JOEDICKE,WILLIAM
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 12/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: BISCONTI

DATE/TIME OF TEST: 01/26/2018 02:08:59

Diagnostics Check		OK	01:57	
Air Blank		0.000	01:58	
Control Test		0.082	01:58	
Air Blank		0.000	01:58	
Subject Sample	# 1		02:02	No Sample Provided
Breath Volume		0.000	02:02	
Air Blank		0.000	02:02	
Air Blank		0.000	02:04	
Subject Sample	# 2		02:07	No Sample Provided
Breath Volume		0.000	02:07	
Air Blank		0.000	02:08	
Control Test		0.078	02:08	
Air Blank		0.000	02:08	
Diagnostics Check		OK	02:08	

Data Download Date/Time: 01/31/2018 12:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005477

INSTRUMENT REGISTERED TO: DADE CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRISCO,RICKY,A
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: CLOVERSETTLE,CHARLES
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 12/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 16:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:26

SUBJECT'S LAST NAME: EMERSON

DATE/TIME OF TEST: 01/01/2018 17:14:22

Diagnostics Check		OK	17:07
Air Blank		0.000	17:07
Control Test		0.080	17:08
Air Blank		0.000	17:08
Subject Sample	# 1	0.224	17:09
Breath Volume		1.929	17:09
Air Blank		0.000	17:10
Air Blank		0.000	17:12
Subject Sample	# 2	0.216	17:12
Breath Volume		2.023	17:12
Air Blank		0.000	17:13
Control Test		0.078	17:13
Air Blank		0.000	17:14
Diagnostics Check		OK	17:14

Data Download Date/Time: 01/17/2018 17:48

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIMA,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: LIMA,MICHAEL
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 12/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: VILLANUEVAPONCE

DATE/TIME OF TEST: 01/01/2018 01:37:03

Diagnostics Check		OK	01:28
Air Blank		0.000	01:29
Control Test		0.080	01:29
Air Blank		0.000	01:29
Subject Sample	# 1	0.205	01:32
Breath Volume		1.828	01:32
Air Blank		0.000	01:33
Air Blank		0.000	01:34
Subject Sample	# 2	0.202	01:35
Breath Volume		2.042	01:35
Air Blank		0.000	01:36
Control Test		0.078	01:36
Air Blank		0.000	01:36
Diagnostics Check		OK	01:37

Data Download Date/Time: 01/28/2018 16:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,MATTHEW,C
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: MANJASEK,ALEXANDER
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 12/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 05:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:35

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 01/01/2018 06:31:40

Diagnostics Check	OK	06:30	
Air Blank	0.000	06:31	
		06:31	RFI Detect
Air Blank		06:31	RFI Detect

Data Download Date/Time: 01/28/2018 16:27

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAMPE,CORTLAND,B
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: ROCES,J
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 12/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: MICHAEL

DATE/TIME OF TEST: 01/04/2018 01:01:10

Diagnostics Check		OK	00:53	
Air Blank		0.000	00:54	
Control Test		0.082	00:54	
Air Blank		0.000	00:54	
Subject Sample	# 1	0.188	00:55	
Breath Volume		3.382	00:55	
Air Blank		0.000	00:56	
Air Blank		0.000	00:58	
Subject Sample	# 2		00:59	Slope Not Met
Breath Volume		0.695	00:59	
Air Blank		0.000	01:00	
Control Test		0.081	01:00	
Air Blank		0.000	01:01	
Diagnostics Check		OK	01:01	

Data Download Date/Time: 01/28/2018 16:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAMPE,CORTLAND,B
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: ROCES,J
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 12/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: MICHAEL

DATE/TIME OF TEST: 01/04/2018 01:11:35

Diagnostics Check		OK	01:04
Air Blank		0.000	01:05
Control Test		0.081	01:05
Air Blank		0.000	01:06
Subject Sample	# 1	0.179	01:06
Breath Volume		2.359	01:06
Air Blank		0.000	01:07
Air Blank		0.000	01:09
Subject Sample	# 2	0.168	01:09
Breath Volume		2.203	01:09
Air Blank		0.000	01:10
Control Test		0.081	01:10
Air Blank		0.000	01:11
Diagnostics Check		OK	01:11

Data Download Date/Time: 01/28/2018 16:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAMPE,CORTLAND,B
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: EDWARDS,L
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 12/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 04:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:01

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 01/14/2018 05:31:15

Diagnostics Check		OK	05:23
Air Blank		0.000	05:24
Control Test		0.081	05:24
Air Blank		0.000	05:25
Subject Sample	# 1	0.191	05:26
Breath Volume		2.449	05:26
Air Blank		0.000	05:27
Air Blank		0.000	05:28
Subject Sample	# 2	0.185	05:29
Breath Volume		2.722	05:29
Air Blank		0.000	05:30
Control Test		0.081	05:30
Air Blank		0.000	05:31
Diagnostics Check		OK	05:31

Data Download Date/Time: 01/28/2018 16:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,JAMES,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: GRAZIANO,ANGELO
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 12/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: WALLACE

DATE/TIME OF TEST: 01/17/2018 04:22:51

Diagnostics Check		OK	04:14
Air Blank		0.000	04:15
Control Test		0.081	04:15
Air Blank		0.000	04:15
Subject Sample	# 1	0.180	04:17
Breath Volume		1.585	04:17
Air Blank		0.000	04:18
Air Blank		0.000	04:20
Subject Sample	# 2	0.194	04:21
Breath Volume		1.644	04:21
Air Blank		0.000	04:21
Control Test		0.079	04:22
Air Blank		0.000	04:22
Diagnostics Check		OK	04:22

Data Download Date/Time: 01/28/2018 16:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,JAMES,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: MCCARTHY,EVAN
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 01/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: DINDYAL

DATE/TIME OF TEST: 01/29/2018 22:57:57

Diagnostics Check		OK	22:51
Air Blank		0.000	22:51
Control Test		0.081	22:52
Air Blank		0.000	22:52
Subject Sample	# 1	0.134	22:53
Breath Volume		1.933	22:53
Air Blank		0.000	22:53
Air Blank		0.000	22:55
Subject Sample	# 2	0.125	22:56
Breath Volume		2.140	22:56
Air Blank		0.000	22:57
Control Test		0.080	22:57
Air Blank		0.000	22:57
Diagnostics Check		OK	22:57

Data Download Date/Time: 02/25/2018 23:16

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAETSCH,JOSH,E
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PAETSCH,JOSH
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 17316080A2
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: SIMPSON

DATE/TIME OF TEST: 01/11/2018 00:27:34

Diagnostics Check		OK	00:20
Air Blank		0.000	00:21
Control Test		0.080	00:21
Air Blank		0.000	00:21
Subject Sample	# 1	0.351	00:22
Breath Volume		2.781	00:22
Air Blank		0.000	00:23
Air Blank		0.000	00:25
Subject Sample	# 2	0.352	00:25
Breath Volume		3.644	00:25
Air Blank		0.000	00:26
Control Test		0.078	00:26
Air Blank		0.000	00:27
Diagnostics Check		OK	00:27

Data Download Date/Time: 02/15/2018 16:43

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,J,
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: MORENO,S
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A2
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: ESTEBAN

DATE/TIME OF TEST: 01/19/2018 02:37:20

Diagnostics Check		OK	02:30
Air Blank		0.000	02:31
Control Test		0.081	02:31
Air Blank		0.000	02:32
Subject Sample	# 1	0.199	02:32
Breath Volume		2.812	02:32
Air Blank		0.000	02:33
Air Blank		0.000	02:35
Subject Sample	# 2	0.184	02:35
Breath Volume		2.843	02:35
Air Blank		0.000	02:36
Control Test		0.080	02:36
Air Blank		0.000	02:37
Diagnostics Check		OK	02:37

Data Download Date/Time: 02/15/2018 16:43

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,J,
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: CAREY,J
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A2
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:46

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 01/20/2018 23:24:35

Diagnostics Check		OK	23:21	
Air Blank		0.000	23:22	
Control Test		0.081	23:22	
Air Blank		0.000	23:23	
Subject Sample	# 1		23:23	Subject Test Refused
Breath Volume		0.000	23:23	
Air Blank		0.000	23:23	
Control Test		0.080	23:23	
Air Blank		0.000	23:24	
Diagnostics Check		OK	23:24	

Data Download Date/Time: 02/15/2018 16:43

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JASON,E
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PAETSCH,JOSHUA
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A2
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: LAY

DATE/TIME OF TEST: 01/27/2018 02:58:24

Diagnostics Check		OK	02:51
Air Blank		0.000	02:51
Control Test		0.081	02:52
Air Blank		0.000	02:52
Subject Sample	# 1	0.182	02:53
Breath Volume		1.648	02:53
Air Blank		0.000	02:54
Air Blank		0.000	02:56
Subject Sample	# 2	0.186	02:56
Breath Volume		1.593	02:56
Air Blank		0.000	02:57
Control Test		0.079	02:57
Air Blank		0.000	02:58
Diagnostics Check		OK	02:58

Data Download Date/Time: 02/15/2018 16:43

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAETSCH,JOSH,E
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PAETSCH,JOSH
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A2
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: BYRNE

DATE/TIME OF TEST: 01/28/2018 03:52:57

Diagnostics Check		OK	03:40
Air Blank		0.000	03:40
Control Test		0.079	03:40
Air Blank		0.000	03:41
Subject Sample	# 1	0.203	03:44
Breath Volume		2.019	03:44
Air Blank		0.000	03:45
Air Blank		0.000	03:47
Subject Sample	# 2	0.182	03:47
Breath Volume		1.289	03:47
Air Blank		0.000	03:48
Air Blank		0.000	03:50
Subject Sample	# 3	0.203	03:51
Breath Volume		2.339	03:51
Air Blank		0.000	03:51
Control Test		0.077	03:52
Air Blank		0.000	03:52
Diagnostics Check		OK	03:52

Data Download Date/Time: 02/15/2018 16:43

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: RUBIO-NUNEZ

DATE/TIME OF TEST: 01/19/2018 23:41:43

Diagnostics Check		OK	23:35
Air Blank		0.000	23:36
Control Test		0.081	23:36
Air Blank		0.000	23:37
Subject Sample	# 1	0.000	23:37
Breath Volume		2.402	23:37
Air Blank		0.000	23:38
Air Blank		0.000	23:40
Subject Sample	# 2	0.000	23:40
Breath Volume		1.445	23:40
Air Blank		0.000	23:40
Control Test		0.079	23:41
Air Blank		0.000	23:41
Diagnostics Check		OK	23:41

Data Download Date/Time: 02/08/2018 09:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEON,URIAK,
BREATH TEST OPERATOR AGENCY: Hialeah Gardens PD
ARRESTING OFFICER: HOSEIN,HASSAN
ARRESTING OFFICER AGENCY: Opa Locka PD
LAST AGENCY INSPECTION: 01/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: ZUNIGA

DATE/TIME OF TEST: 01/29/2018 19:57:41

Diagnostics Check		OK	19:44	
Air Blank		0.000	19:45	
Control Test		0.079	19:45	
Air Blank		0.000	19:46	
Subject Sample	# 1	0.215	19:49	Volume Not Met
Breath Volume		0.800	19:49	
Air Blank		0.000	19:50	
Air Blank		0.000	19:51	
Subject Sample	# 2	0.244	19:53	
Breath Volume		1.179	19:53	
Air Blank		0.000	19:53	
Air Blank		0.000	19:55	
Subject Sample	# 3		19:56	Subject Test Refused
Breath Volume		0.105	19:56	
Air Blank			19:57	RFI Detect
Air Blank		0.000	19:57	

Data Download Date/Time: 02/08/2018 09:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PACHECO,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: SANCHEZ,JIMMY
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 12/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: OCONNER

DATE/TIME OF TEST: 01/08/2018 18:56:05

Diagnostics Check		OK	18:40	
Air Blank		0.000	18:41	
Control Test		0.079	18:41	
Air Blank		0.000	18:42	
Subject Sample	# 1	0.235	18:44	
Breath Volume		1.210	18:44	
Air Blank		0.000	18:45	
Air Blank		0.000	18:46	
Subject Sample	# 2	0.223	18:49	Volume Not Met
Breath Volume		0.160	18:49	
Air Blank		0.000	18:50	
Air Blank		0.000	18:52	
Subject Sample	# 3	0.231	18:54	
Breath Volume		1.167	18:54	
Air Blank		0.000	18:55	
Control Test		0.076	18:55	
Air Blank		0.000	18:56	
Diagnostics Check		OK	18:56	

Data Download Date/Time: 01/24/2018 10:08

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PACHECO,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: SANCHEZ,JIMY
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 12/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: OCONNER

DATE/TIME OF TEST: 01/08/2018 19:13:08

Diagnostics Check		OK	19:06
Air Blank		0.000	19:06
Control Test		0.078	19:07
Air Blank		0.000	19:07
Subject Sample	# 1	0.236	19:08
Breath Volume		1.671	19:08
Air Blank		0.000	19:09
Air Blank		0.000	19:10
Subject Sample	# 2	0.225	19:11
Breath Volume		1.523	19:11
Air Blank		0.000	19:12
Control Test		0.078	19:12
Air Blank		0.000	19:13
Diagnostics Check		OK	19:13

Data Download Date/Time: 01/24/2018 10:08

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONILLA,JOSE,I
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PORTILLO,BRYAN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: VILLA ELIAS

DATE/TIME OF TEST: 01/27/2018 03:23:42

Diagnostics Check		OK	03:20	
Air Blank		0.000	03:21	
Control Test		0.079	03:21	
Air Blank		0.000	03:22	
Subject Sample	# 1		03:22	Subject Test Refused
Breath Volume		0.000	03:22	
Air Blank		0.000	03:22	
Control Test		0.079	03:23	
Air Blank		0.000	03:23	
Diagnostics Check		OK	03:23	

Data Download Date/Time: 02/19/2018 10:53

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONILLA,JOSE,I
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: BONILLA,JOSE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 16:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:27

SUBJECT'S LAST NAME: MANUEL MENDOZA

DATE/TIME OF TEST: 01/27/2018 17:51:15

Diagnostics Check		OK	17:48	
Air Blank		0.000	17:48	
Control Test		0.079	17:49	
Air Blank		0.000	17:49	
Subject Sample	# 1		17:49	Subject Test Refused
Breath Volume		0.000	17:49	
Air Blank		0.000	17:50	
Control Test		0.080	17:50	
Air Blank		0.000	17:51	
Diagnostics Check		OK	17:51	

Data Download Date/Time: 02/19/2018 10:53

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:25

SUBJECT'S LAST NAME: RESTREPO P

DATE/TIME OF TEST: 01/06/2018 05:53:28

Diagnostics Check		OK	05:46
Air Blank		0.000	05:46
Control Test		0.081	05:47
Air Blank		0.000	05:47
Subject Sample	# 1	0.149	05:48
Breath Volume		2.847	05:48
Air Blank		0.000	05:48
Air Blank		0.000	05:50
Subject Sample	# 2	0.150	05:51
Breath Volume		2.855	05:51
Air Blank		0.000	05:52
Control Test		0.081	05:52
Air Blank		0.000	05:53
Diagnostics Check		OK	05:53

Data Download Date/Time: 02/08/2018 11:51

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: ROSELLO

DATE/TIME OF TEST: 01/10/2018 00:46:50

Diagnostics Check		OK	00:38
Air Blank		0.000	00:39
Control Test		0.079	00:39
Air Blank		0.000	00:40
Subject Sample	# 1	0.178	00:41
Breath Volume		1.195	00:41
Air Blank		0.000	00:41
Air Blank		0.000	00:43
Subject Sample	# 2	0.169	00:45
Breath Volume		2.410	00:45
Air Blank		0.000	00:45
Control Test		0.077	00:46
Air Blank		0.000	00:46
Diagnostics Check		OK	00:46

Data Download Date/Time: 02/08/2018 11:51

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: MONTENEGRO

DATE/TIME OF TEST: 01/14/2018 03:28:09

Diagnostics Check		OK	03:18
Air Blank		0.000	03:18
Control Test		0.081	03:18
Air Blank		0.000	03:19
Subject Sample	# 1	0.121	03:22
Breath Volume		1.121	03:22
Air Blank		0.000	03:23
Air Blank		0.000	03:24
Subject Sample	# 2	0.125	03:26
Breath Volume		1.144	03:26
Air Blank		0.000	03:27
Control Test		0.078	03:27
Air Blank		0.000	03:28
Diagnostics Check		OK	03:28

Data Download Date/Time: 02/08/2018 11:51

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: MEMBRENO

DATE/TIME OF TEST: 01/31/2018 03:13:47

Diagnostics Check		OK	03:06
Air Blank		0.000	03:06
Control Test		0.081	03:06
Air Blank		0.000	03:07
Subject Sample	# 1	0.181	03:08
Breath Volume		1.156	03:08
Air Blank		0.000	03:09
Air Blank		0.000	03:11
Subject Sample	# 2	0.199	03:12
Breath Volume		2.046	03:12
Air Blank		0.000	03:12
Control Test		0.081	03:13
Air Blank		0.000	03:13
Diagnostics Check		OK	03:13

Data Download Date/Time: 02/08/2018 11:51

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUFAHL,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PUFAHL,G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: OROZCO VASQUEZ

DATE/TIME OF TEST: 01/04/2018 00:26:03

Diagnostics Check		OK	00:19
Air Blank		0.000	00:20
Control Test		0.082	00:20
Air Blank		0.000	00:20
Subject Sample	# 1	0.205	00:21
Breath Volume		2.476	00:21
Air Blank		0.000	00:22
Air Blank		0.000	00:23
Subject Sample	# 2	0.208	00:24
Breath Volume		2.234	00:24
Air Blank		0.000	00:25
Control Test		0.081	00:25
Air Blank		0.000	00:25
Diagnostics Check		OK	00:26

Data Download Date/Time: 01/31/2018 09:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUFAHL,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PUFAHL,G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: HEELEY

DATE/TIME OF TEST: 01/07/2018 02:33:09

Diagnostics Check		OK	02:26
Air Blank		0.000	02:27
Control Test		0.082	02:27
Air Blank		0.000	02:27
Subject Sample	# 1	0.220	02:28
Breath Volume		4.652	02:28
Air Blank		0.000	02:29
Air Blank		0.000	02:30
Subject Sample	# 2	0.221	02:31
Breath Volume		3.386	02:31
Air Blank		0.000	02:32
Control Test		0.081	02:32
Air Blank		0.000	02:33
Diagnostics Check		OK	02:33

Data Download Date/Time: 01/31/2018 09:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUFAHL,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PUFAHL,G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/29/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: CAMACHO ORTIZ

DATE/TIME OF TEST: 01/11/2018 01:06:11

Diagnostics Check		OK	00:59
Air Blank		0.000	01:00
Control Test		0.079	01:00
Air Blank		0.000	01:00
Subject Sample	# 1	0.251	01:01
Breath Volume		1.613	01:01
Air Blank		0.000	01:02
Air Blank		0.000	01:03
Subject Sample	# 2	0.262	01:04
Breath Volume		2.460	01:04
Air Blank		0.000	01:05
Control Test		0.080	01:05
Air Blank		0.000	01:06
Diagnostics Check		OK	01:06

Data Download Date/Time: 01/31/2018 09:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUFAHL,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PUFAHL,G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: RICAUD

DATE/TIME OF TEST: 01/19/2018 02:05:35

Diagnostics Check		OK	01:56
Air Blank		0.000	01:57
Control Test		0.082	01:57
Air Blank		0.000	01:58
Subject Sample	# 1	0.183	02:00
Breath Volume		2.023	02:00
Air Blank		0.000	02:01
Air Blank		0.000	02:02
Subject Sample	# 2	0.197	02:03
Breath Volume		2.160	02:03
Air Blank		0.000	02:04
Control Test		0.076	02:04
Air Blank		0.000	02:05
Diagnostics Check		OK	02:05

Data Download Date/Time: 01/31/2018 09:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUFAHL,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PUFAHL,G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/29/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 01/19/2018 22:20:40

Diagnostics Check		OK	22:14
Air Blank		0.000	22:14
Control Test		0.081	22:14
Air Blank		0.000	22:15
Subject Sample	# 1	0.233	22:16
Breath Volume		2.125	22:16
Air Blank		0.000	22:16
Air Blank		0.000	22:18
Subject Sample	# 2	0.224	22:19
Breath Volume		2.035	22:19
Air Blank		0.000	22:19
Control Test		0.079	22:20
Air Blank		0.000	22:20
Diagnostics Check		OK	22:20

Data Download Date/Time: 01/31/2018 09:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUFAHL,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PUFAHL,G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: CONNETT

DATE/TIME OF TEST: 01/20/2018 03:00:39

Diagnostics Check		OK	02:54
Air Blank		0.000	02:54
Control Test		0.081	02:54
Air Blank		0.000	02:55
Subject Sample	# 1	0.260	02:55
Breath Volume		2.847	02:55
Air Blank		0.000	02:56
Air Blank		0.000	02:58
Subject Sample	# 2	0.258	02:58
Breath Volume		2.656	02:58
Air Blank		0.000	02:59
Control Test		0.079	03:00
Air Blank		0.000	03:00
Diagnostics Check		OK	03:00

Data Download Date/Time: 01/31/2018 09:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUFAHL,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PUFAHL,G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: GRESS

DATE/TIME OF TEST: 01/24/2018 22:28:36

Diagnostics Check		OK	22:17	
Air Blank		0.000	22:17	
Control Test		0.081	22:17	
Air Blank		0.000	22:18	
Subject Sample	# 1		22:21	No Sample Provided
Breath Volume		0.000	22:21	
Air Blank		0.000	22:22	
Air Blank		0.000	22:24	
Subject Sample	# 2		22:27	No Sample Provided
Breath Volume		0.000	22:27	
Air Blank		0.000	22:27	
Control Test		0.081	22:27	
Air Blank		0.000	22:28	
Diagnostics Check		OK	22:28	

Data Download Date/Time: 01/31/2018 09:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FICARRA,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: CUE

DATE/TIME OF TEST: 01/01/2018 02:19:30

Diagnostics Check		OK	02:12
Air Blank		0.000	02:13
Control Test		0.081	02:13
Air Blank		0.000	02:14
Subject Sample	# 1	0.193	02:14
Breath Volume		2.589	02:14
Air Blank		0.000	02:15
Air Blank		0.000	02:17
Subject Sample	# 2	0.197	02:17
Breath Volume		2.339	02:17
Air Blank		0.000	02:18
Control Test		0.079	02:18
Air Blank		0.000	02:19
Diagnostics Check		OK	02:19

Data Download Date/Time: 01/19/2018 16:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NOEL,KENNETH
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 01/01/2018 03:27:17

Diagnostics Check		OK	03:21	
Air Blank		0.000	03:22	
Control Test		0.081	03:22	
Air Blank		0.000	03:22	
Subject Sample	# 1		03:25	Subject Test Refused
Breath Volume		0.000	03:25	
Air Blank		0.000	03:26	
Control Test		0.081	03:26	
Air Blank		0.000	03:27	
Diagnostics Check		OK	03:27	

Data Download Date/Time: 01/19/2018 16:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LEHENY,JACOB
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: HIGHTOWER

DATE/TIME OF TEST: 01/01/2018 04:01:49

Diagnostics Check		OK	03:53	
Air Blank		0.000	03:53	
Control Test		0.081	03:54	
Air Blank		0.000	03:54	
Subject Sample	# 1	0.306	03:57	Volume Not Met
Breath Volume		0.214	03:57	
Air Blank		0.000	03:58	
Air Blank		0.000	04:00	
Subject Sample	# 2		04:00	Subject Test Refused
Breath Volume		0.000	04:00	
Air Blank		0.000	04:00	
Control Test		0.076	04:01	
Air Blank		0.000	04:01	
Diagnostics Check		OK	04:01	

Data Download Date/Time: 01/19/2018 16:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: YOCHUM,CRAIG
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 01/01/2018 04:31:41

Diagnostics Check		OK	04:25
Air Blank		0.000	04:25
Control Test		0.080	04:26
Air Blank		0.000	04:26
Subject Sample	# 1	0.171	04:27
Breath Volume		3.468	04:27
Air Blank		0.000	04:27
Air Blank		0.000	04:29
Subject Sample	# 2	0.174	04:30
Breath Volume		2.562	04:30
Air Blank		0.000	04:30
Control Test		0.080	04:31
Air Blank		0.000	04:31
Diagnostics Check		OK	04:31

Data Download Date/Time: 01/19/2018 16:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LEVEY,SCOTT
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: LA ROSA

DATE/TIME OF TEST: 01/01/2018 05:29:27

Diagnostics Check		OK	05:22
Air Blank		0.000	05:23
Control Test		0.081	05:23
Air Blank		0.000	05:24
Subject Sample	# 1	0.119	05:24
Breath Volume		4.117	05:24
Air Blank		0.000	05:25
Air Blank		0.000	05:27
Subject Sample	# 2	0.120	05:27
Breath Volume		4.800	05:27
Air Blank		0.000	05:28
Control Test		0.080	05:28
Air Blank		0.000	05:29
Diagnostics Check		OK	05:29

Data Download Date/Time: 01/19/2018 16:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHNEIDER,JOSHUA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 01/04/2018 23:59:45

Diagnostics Check		OK	23:52
Air Blank		0.000	23:52
Control Test		0.083	23:53
Air Blank		0.000	23:53
Subject Sample	# 1	0.060	23:55
Breath Volume		1.519	23:55
Air Blank		0.000	23:55
Air Blank		0.000	23:57
Subject Sample	# 2	0.062	23:58
Breath Volume		1.636	23:58
Air Blank		0.000	23:58
Control Test		0.081	23:59
Air Blank		0.000	23:59
Diagnostics Check		OK	23:59

Data Download Date/Time: 01/19/2018 16:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHNEIDER,JOSHUA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: CALDERA

DATE/TIME OF TEST: 01/05/2018 03:20:41

Diagnostics Check		OK	03:13
Air Blank		0.000	03:14
Control Test		0.082	03:14
Air Blank		0.000	03:15
Subject Sample	# 1	0.240	03:15
Breath Volume		1.968	03:15
Air Blank		0.000	03:16
Air Blank		0.000	03:18
Subject Sample	# 2	0.234	03:19
Breath Volume		1.289	03:19
Air Blank		0.000	03:19
Control Test		0.082	03:20
Air Blank		0.000	03:20
Diagnostics Check		OK	03:20

Data Download Date/Time: 01/19/2018 16:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HANTON,MELINDA
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: MANSUR

DATE/TIME OF TEST: 01/06/2018 00:21:20

Diagnostics Check		OK	00:08	
Air Blank		0.000	00:08	
Control Test		0.081	00:09	
Air Blank		0.000	00:09	
Subject Sample	# 1	0.201	00:10	
Breath Volume		1.105	00:10	
Air Blank		0.000	00:11	
Air Blank		0.000	00:13	
Subject Sample	# 2	0.172	00:16	Volume Not Met
Breath Volume		0.832	00:16	
Air Blank		0.000	00:17	
Air Blank		0.000	00:18	
Subject Sample	# 3	0.198	00:19	
Breath Volume		2.062	00:19	
Air Blank		0.000	00:20	
Control Test		0.077	00:20	
Air Blank		0.000	00:21	
Diagnostics Check		OK	00:21	

Data Download Date/Time: 01/19/2018 16:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: SOSA MARTINEZ

DATE/TIME OF TEST: 01/06/2018 01:24:27

Diagnostics Check		OK	01:11
Air Blank		0.000	01:11
Control Test		0.081	01:12
Air Blank		0.000	01:12
Subject Sample	# 1	0.141	01:14
Breath Volume		1.582	01:14
Air Blank		0.000	01:15
Air Blank		0.000	01:17
Subject Sample	# 2	0.119	01:17
Breath Volume		1.160	01:17
Air Blank		0.000	01:18
Air Blank		0.000	01:20
Subject Sample	# 3	0.124	01:22
Breath Volume		1.250	01:22
Air Blank		0.000	01:23
Control Test		0.077	01:23
Air Blank		0.000	01:24
Diagnostics Check		OK	01:24

Data Download Date/Time: 01/19/2018 16:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TAVARES,LUCAS
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: VANDERLOO

DATE/TIME OF TEST: 01/07/2018 01:10:48

Diagnostics Check		OK	01:03
Air Blank		0.000	01:04
Control Test		0.081	01:04
Air Blank		0.000	01:05
Subject Sample	# 1	0.191	01:06
Breath Volume		2.925	01:06
Air Blank		0.000	01:07
Air Blank		0.000	01:08
Subject Sample	# 2	0.173	01:09
Breath Volume		2.140	01:09
Air Blank		0.000	01:09
Control Test		0.079	01:10
Air Blank		0.000	01:10
Diagnostics Check		OK	01:10

Data Download Date/Time: 01/19/2018 16:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LEVEY,SCOTT
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: CARDARELLI

DATE/TIME OF TEST: 01/07/2018 02:12:05

Diagnostics Check		OK	02:04
Air Blank		0.000	02:05
Control Test		0.081	02:05
Air Blank		0.000	02:05
Subject Sample	# 1	0.080	02:07
Breath Volume		1.601	02:07
Air Blank		0.000	02:08
Air Blank		0.000	02:10
Subject Sample	# 2	0.083	02:10
Breath Volume		1.941	02:10
Air Blank		0.000	02:11
Control Test		0.079	02:11
Air Blank		0.000	02:12
Diagnostics Check		OK	02:12

Data Download Date/Time: 01/19/2018 16:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: RICCIARDI,FRANK
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: ESCALANTE JIMENEZ

DATE/TIME OF TEST: 01/07/2018 03:59:24

Diagnostics Check		OK	03:51
Air Blank		0.000	03:51
Control Test		0.081	03:52
Air Blank		0.000	03:52
Subject Sample	# 1	0.220	03:53
Breath Volume		1.882	03:53
Air Blank		0.000	03:54
Air Blank		0.000	03:56
Subject Sample	# 2	0.206	03:57
Breath Volume		1.714	03:57
Air Blank		0.000	03:58
Control Test		0.078	03:58
Air Blank		0.000	03:59
Diagnostics Check		OK	03:59

Data Download Date/Time: 01/19/2018 16:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCARTOZZI,PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:33

SUBJECT'S LAST NAME: KIMANI

DATE/TIME OF TEST: 01/07/2018 05:07:20

Diagnostics Check		OK	05:00
Air Blank		0.000	05:00
Control Test		0.080	05:00
Air Blank		0.000	05:01
Subject Sample	# 1	0.263	05:02
Breath Volume		1.175	05:02
Air Blank		0.000	05:03
Air Blank		0.000	05:05
Subject Sample	# 2	0.261	05:05
Breath Volume		2.167	05:05
Air Blank		0.000	05:06
Control Test		0.079	05:06
Air Blank		0.000	05:07
Diagnostics Check		OK	05:07

Data Download Date/Time: 01/19/2018 16:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GARRETT,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: SMITH JR

DATE/TIME OF TEST: 01/07/2018 20:20:13

Diagnostics Check		OK	20:13
Air Blank		0.000	20:14
Control Test		0.081	20:14
Air Blank		0.000	20:15
Subject Sample	# 1	0.193	20:15
Breath Volume		1.261	20:15
Air Blank		0.000	20:16
Air Blank		0.000	20:18
Subject Sample	# 2	0.199	20:18
Breath Volume		1.667	20:18
Air Blank		0.000	20:19
Control Test		0.081	20:19
Air Blank		0.000	20:20
Diagnostics Check		OK	20:20

Data Download Date/Time: 01/19/2018 16:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: BARCELO NOGUEIRA

DATE/TIME OF TEST: 01/09/2018 00:11:07

Diagnostics Check		OK	00:04
Air Blank		0.000	00:05
Control Test		0.082	00:05
Air Blank		0.000	00:05
Subject Sample	# 1	0.215	00:06
Breath Volume		1.781	00:06
Air Blank		0.000	00:07
Air Blank		0.000	00:09
Subject Sample	# 2	0.212	00:09
Breath Volume		2.644	00:09
Air Blank		0.000	00:10
Control Test		0.080	00:10
Air Blank		0.000	00:11
Diagnostics Check		OK	00:11

Data Download Date/Time: 01/19/2018 16:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FREY,JACOB
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: MCARDLE

DATE/TIME OF TEST: 01/09/2018 22:19:35

Diagnostics Check		OK	22:13
Air Blank		0.000	22:13
Control Test		0.081	22:13
Air Blank		0.000	22:14
Subject Sample	# 1	0.186	22:15
Breath Volume		2.652	22:15
Air Blank		0.000	22:15
Air Blank		0.000	22:17
Subject Sample	# 2	0.187	22:17
Breath Volume		2.875	22:17
Air Blank		0.000	22:18
Control Test		0.080	22:18
Air Blank		0.000	22:19
Diagnostics Check		OK	22:19

Data Download Date/Time: 01/19/2018 16:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SORIANO,REYNALDO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: GILBERT

DATE/TIME OF TEST: 01/11/2018 00:07:33

Diagnostics Check		OK	23:59	
Air Blank		0.000	23:59	
Control Test		0.081	23:59	
Air Blank		0.000	00:00	
Subject Sample	# 1	0.154	00:03	Volume Not Met
Breath Volume		0.082	00:03	
Air Blank		0.000	00:04	
Air Blank		0.000	00:05	
Subject Sample	# 2		00:06	Subject Test Refused
Breath Volume		0.000	00:06	
Air Blank		0.000	00:06	
Control Test		0.079	00:06	
Air Blank		0.000	00:07	
Diagnostics Check		OK	00:07	

Data Download Date/Time: 01/19/2018 16:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHNEIDER,JOSHUA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: SAMUELS

DATE/TIME OF TEST: 01/12/2018 00:03:12

Diagnostics Check		OK	23:57
Air Blank		0.000	23:57
Control Test		0.081	23:57
Air Blank		0.000	23:58
Subject Sample	# 2	0.000	23:58
Breath Volume		1.117	23:58
Air Blank		0.000	23:59
Air Blank		0.000	00:01
Subject Sample	# 1	0.000	00:01
Breath Volume		1.417	00:01
Air Blank		0.000	00:02
Control Test		0.081	00:02
Air Blank		0.000	00:03
Diagnostics Check		OK	00:03

Data Download Date/Time: 01/19/2018 16:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TAVARES,LUCAS
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: TRUBCHIKOV

DATE/TIME OF TEST: 01/12/2018 02:51:04

Diagnostics Check		OK	02:39
Air Blank		0.000	02:39
Control Test		0.082	02:40
Air Blank		0.000	02:40
Subject Sample	# 1	0.173	02:42
Breath Volume		3.304	02:42
Air Blank		0.000	02:43
Air Blank		0.000	02:44
Subject Sample	# 2	0.149	02:45
Breath Volume		1.531	02:45
Air Blank		0.000	02:46
Air Blank		0.000	02:48
Subject Sample	# 3	0.147	02:49
Breath Volume		1.757	02:49
Air Blank		0.000	02:50
Control Test		0.079	02:50
Air Blank		0.000	02:50
Diagnostics Check		OK	02:51

Data Download Date/Time: 01/19/2018 16:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOYNIHAN,KERIANE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHNEIDER,JOSHUA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:27

SUBJECT'S LAST NAME: KERN

DATE/TIME OF TEST: 01/12/2018 19:02:06

Diagnostics Check		OK	18:55
Air Blank		0.000	18:55
Control Test		0.082	18:55
Air Blank		0.000	18:56
Subject Sample	# 1	0.202	18:57
Breath Volume		2.273	18:57
Air Blank		0.000	18:57
Air Blank		0.000	18:59
Subject Sample	# 2	0.206	19:00
Breath Volume		2.121	19:00
Air Blank		0.000	19:01
Control Test		0.079	19:01
Air Blank		0.000	19:02
Diagnostics Check		OK	19:02

Data Download Date/Time: 01/19/2018 16:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOYNIHAN,KERIANE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CHIQUITO-RODRIGUEZ,LENNDY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Male

OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: MOLLOY

DATE/TIME OF TEST: 01/12/2018 23:48:33

Diagnostics Check		OK	23:41
Air Blank		0.000	23:42
Control Test		0.081	23:42
Air Blank		0.000	23:43
Subject Sample	# 1	0.311	23:43
Breath Volume		2.046	23:43
Air Blank		0.000	23:44
Air Blank		0.000	23:46
Subject Sample	# 2	0.324	23:46
Breath Volume		2.820	23:46
Air Blank		0.000	23:47
Control Test		0.080	23:47
Air Blank		0.000	23:48
Diagnostics Check		OK	23:48

Data Download Date/Time: 01/19/2018 16:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TODD,ZION
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: PEREZ MORALES

DATE/TIME OF TEST: 01/13/2018 04:11:03

Diagnostics Check		OK	04:04
Air Blank		0.000	04:05
Control Test		0.081	04:05
Air Blank		0.000	04:05
Subject Sample	# 1	0.170	04:06
Breath Volume		1.148	04:06
Air Blank		0.000	04:07
Air Blank		0.000	04:08
Subject Sample	# 2	0.172	04:09
Breath Volume		1.296	04:09
Air Blank		0.000	04:10
Control Test		0.080	04:10
Air Blank		0.000	04:10
Diagnostics Check		OK	04:11

Data Download Date/Time: 01/19/2018 16:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOYNIHAN,KERIANE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LEVEY,SCOTT
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 15:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:11

SUBJECT'S LAST NAME: FOUSNOUGHT

DATE/TIME OF TEST: 01/13/2018 16:44:42

Diagnostics Check		OK	16:37
Air Blank		0.000	16:38
Control Test		0.081	16:38
Air Blank		0.000	16:39
Subject Sample	# 1	0.138	16:39
Breath Volume		3.437	16:39
Air Blank		0.000	16:40
Air Blank		0.000	16:42
Subject Sample	# 2	0.131	16:42
Breath Volume		3.027	16:42
Air Blank		0.000	16:43
Control Test		0.080	16:44
Air Blank		0.000	16:44
Diagnostics Check		OK	16:44

Data Download Date/Time: 01/19/2018 16:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOYNIHAN,KERIANE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LAMBERT,ANWAR
ARRESTING OFFICER AGENCY: Riviera Beach PD
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:01

SUBJECT'S LAST NAME: GRIGGS

DATE/TIME OF TEST: 01/13/2018 21:43:08

Diagnostics Check		OK	21:40	
Air Blank		0.000	21:40	
Control Test		0.081	21:41	
Air Blank		0.000	21:41	
Subject Sample	# 1		21:42	Sequence Aborted
Breath Volume		0.000	21:42	
Air Blank		0.000	21:43	

Data Download Date/Time: 01/19/2018 16:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOYNIHAN,KERIANE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: RAMIREZ,DAYRON
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: MEJIA

DATE/TIME OF TEST: 01/14/2018 00:54:50

Diagnostics Check		OK	00:47
Air Blank		0.000	00:48
Control Test		0.081	00:48
Air Blank		0.000	00:49
Subject Sample	# 1	0.201	00:50
Breath Volume		3.027	00:50
Air Blank		0.000	00:50
Air Blank		0.000	00:52
Subject Sample	# 2	0.201	00:53
Breath Volume		1.757	00:53
Air Blank		0.000	00:53
Control Test		0.080	00:54
Air Blank		0.000	00:54
Diagnostics Check		OK	00:54

Data Download Date/Time: 01/19/2018 16:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SALVEMINI,MICHAEL
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 01/14/2018 02:46:20

Diagnostics Check		OK	02:38
Air Blank		0.000	02:39
Control Test		0.082	02:39
Air Blank		0.000	02:40
Subject Sample	# 1	0.247	02:41
Breath Volume		3.199	02:41
Air Blank		0.000	02:42
Air Blank		0.000	02:43
Subject Sample	# 2	0.238	02:44
Breath Volume		3.101	02:44
Air Blank		0.000	02:45
Control Test		0.079	02:45
Air Blank		0.000	02:46
Diagnostics Check		OK	02:46

Data Download Date/Time: 01/19/2018 16:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CHIQUITO-RODRIGUEZ,LENNDY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:17
GENDER: Male

OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 01/14/2018 21:42:42

Diagnostics Check		OK	21:35
Air Blank		0.000	21:36
Control Test		0.082	21:36
Air Blank		0.000	21:37
Subject Sample	# 1	0.189	21:37
Breath Volume		4.488	21:37
Air Blank		0.000	21:38
Air Blank		0.000	21:40
Subject Sample	# 2	0.187	21:40
Breath Volume		4.023	21:40
Air Blank		0.000	21:41
Control Test		0.080	21:42
Air Blank		0.000	21:42
Diagnostics Check		OK	21:42

Data Download Date/Time: 01/19/2018 16:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: AMAYA BONILLA

DATE/TIME OF TEST: 01/15/2018 00:29:19

Diagnostics Check		OK	00:22
Air Blank		0.000	00:23
Control Test		0.081	00:23
Air Blank		0.000	00:24
Subject Sample	# 1	0.236	00:24
Breath Volume		3.496	00:24
Air Blank		0.000	00:25
Air Blank		0.000	00:27
Subject Sample	# 2	0.230	00:27
Breath Volume		3.488	00:27
Air Blank		0.000	00:28
Control Test		0.081	00:28
Air Blank		0.000	00:29
Diagnostics Check		OK	00:29

Data Download Date/Time: 01/19/2018 16:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BORROWS,ANDREW
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: PROPPER

DATE/TIME OF TEST: 01/17/2018 00:33:57

Diagnostics Check		OK	00:27
Air Blank		0.000	00:28
Control Test		0.081	00:28
Air Blank		0.000	00:29
Subject Sample	# 1	0.084	00:29
Breath Volume		1.886	00:29
Air Blank		0.000	00:30
Air Blank		0.000	00:31
Subject Sample	# 2	0.085	00:32
Breath Volume		1.718	00:32
Air Blank		0.000	00:33
Control Test		0.080	00:33
Air Blank		0.000	00:33
Diagnostics Check		OK	00:33

Data Download Date/Time: 01/19/2018 16:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOYNIHAN,KERIANE,M
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: TAVARES,LUCAS
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:43

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 01/17/2018 22:17:10

Diagnostics Check		OK	22:10
Air Blank		0.000	22:10
Control Test		0.081	22:10
Air Blank		0.000	22:11
Subject Sample	# 1	0.200	22:12
Breath Volume		2.082	22:12
Air Blank		0.000	22:12
Air Blank		0.000	22:14
Subject Sample	# 2	0.203	22:15
Breath Volume		2.816	22:15
Air Blank		0.000	22:16
Control Test		0.082	22:16
Air Blank		0.000	22:17
Diagnostics Check		OK	22:17

Data Download Date/Time: 01/19/2018 16:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BORROWS,ANDREW
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: BRACCHI

DATE/TIME OF TEST: 01/18/2018 05:35:43

Diagnostics Check		OK	05:27
Air Blank		0.000	05:28
Control Test		0.082	05:28
Air Blank		0.000	05:29
Subject Sample	# 1	0.195	05:30
Breath Volume		2.957	05:30
Air Blank		0.000	05:31
Air Blank		0.000	05:32
Subject Sample	# 2	0.197	05:33
Breath Volume		3.187	05:33
Air Blank		0.000	05:34
Control Test		0.080	05:35
Air Blank		0.000	05:35
Diagnostics Check		OK	05:35

Data Download Date/Time: 01/19/2018 16:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOYNIHAN,KERIANE,M
BREATH TEST OPERATOR AGENCY: Jupiter PD
ARRESTING OFFICER: YOCHUM,CRAIG
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: DUKE

DATE/TIME OF TEST: 01/18/2018 23:36:33

Diagnostics Check		OK	23:30
Air Blank		0.000	23:30
Control Test		0.082	23:30
Air Blank		0.000	23:31
Subject Sample	# 1	0.020	23:32
Breath Volume		1.457	23:32
Air Blank		0.000	23:32
Air Blank		0.000	23:34
Subject Sample	# 2	0.020	23:35
Breath Volume		1.144	23:35
Air Blank		0.000	23:35
Control Test		0.082	23:35
Air Blank		0.000	23:36
Diagnostics Check		OK	23:36

Data Download Date/Time: 01/19/2018 16:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006046

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,DANIELLE,M
BREATH TEST OPERATOR AGENCY: FDLE
ARRESTING OFFICER: BELL,DANIELLE
ARRESTING OFFICER AGENCY: FDLE
LAST AGENCY INSPECTION: 01/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG626604
DRY GAS EXPIRY DATE: 09/22/2018
VIOLATION CODE: SYS CHECK
ARREST TIME: 10:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:30

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 01/05/2018 11:05:59

Diagnostics Check		OK	11:00
Air Blank		0.000	11:00
Control Test		0.079	11:01
Air Blank		0.000	11:01
Subject Sample	# 1	0.000	11:01
Breath Volume		1.359	11:01
Air Blank		0.000	11:02
Air Blank		0.000	11:04
Subject Sample	# 2	0.000	11:04
Breath Volume		1.648	11:04
Air Blank		0.000	11:05
Control Test		0.079	11:05
Air Blank		0.000	11:05
Diagnostics Check		OK	11:05

Data Download Date/Time: 01/08/2018 14:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006046

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALFY,RACHEL,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: PALFY,RACHEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 20:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:29

SUBJECT'S LAST NAME: GIBBS

DATE/TIME OF TEST: 01/31/2018 21:01:43

Diagnostics Check		OK	20:54
Air Blank		0.000	20:55
Control Test		0.079	20:55
Air Blank		0.000	20:55
Subject Sample	# 1	0.119	20:57
Breath Volume		1.800	20:57
Air Blank		0.000	20:57
Air Blank		0.000	20:59
Subject Sample	# 2	0.121	21:00
Breath Volume		2.117	21:00
Air Blank		0.000	21:00
Control Test		0.078	21:01
Air Blank		0.000	21:01
Diagnostics Check		OK	21:01

Data Download Date/Time: 02/27/2018 23:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.