

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RADO,RYAN,R  
BREATH TEST OPERATOR AGENCY: St. Augustine PD  
ARRESTING OFFICER: RADO,RYAN  
ARRESTING OFFICER AGENCY: St. Augustine PD  
LAST AGENCY INSPECTION: 03/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 640976  
DRY GAS EXPIRY DATE: 11/21/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: ROBERSON

DATE/TIME OF TEST: 04/01/2018 01:35:42

---

Diagnostics Check	OK	01:27
Air Blank	0.000	01:28
Control Test	0.080	01:28
Air Blank	0.000	01:29
Subject Sample # 1	0.138	01:30
Breath Volume	1.320	01:30
Air Blank	0.000	01:31
Air Blank	0.000	01:32
Subject Sample # 2	0.130	01:34
Breath Volume	1.417	01:34
Air Blank	0.000	01:34
Control Test	0.078	01:35
Air Blank	0.000	01:35
Diagnostics Check	OK	01:35

---

Data Download Date/Time: 04/04/2018 13:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ACKERMAN,JUSTIN,T  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: HITCH,ANDREW  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 03/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 640976  
DRY GAS EXPIRY DATE: 11/21/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: DUGGER

DATE/TIME OF TEST: 04/01/2018 04:23:28

---

Diagnostics Check	OK	04:16
Air Blank	0.000	04:17
Control Test	0.080	04:17
Air Blank	0.000	04:18
Subject Sample # 1	0.136	04:18
Breath Volume	1.558	04:18
Air Blank	0.000	04:19
Air Blank	0.000	04:21
Subject Sample # 2	0.128	04:21
Breath Volume	1.148	04:21
Air Blank	0.000	04:22
Control Test	0.078	04:22
Air Blank	0.000	04:23
Diagnostics Check	OK	04:23

---

Data Download Date/Time: 04/04/2018 13:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DASILVA,ANDRE,F  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: DASILVA,ANDRE  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 03/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 640976  
DRY GAS EXPIRY DATE: 11/21/2019  
VIOLATION CODE: DUI  
ARREST TIME: 16:59  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:47

SUBJECT'S LAST NAME: GONCALVES

DATE/TIME OF TEST: 04/01/2018 19:25:13

---

Diagnostics Check	OK	19:09	
Air Blank	0.000	19:10	
Control Test	0.080	19:10	
Air Blank	0.000	19:11	
Subject Sample # 1	0.165	19:12	
Breath Volume	1.718	19:12	
Air Blank	0.000	19:13	
Air Blank	0.000	19:14	
Subject Sample # 2	0.139	19:17	Slope Not Level
Breath Volume	1.417	19:17	
Air Blank	0.000	19:18	
Air Blank	0.000	19:20	
Subject Sample # 3	0.135	19:23	Volume Not Met
Breath Volume	0.929	19:23	
Air Blank	0.000	19:24	
Control Test	0.078	19:24	
Air Blank	0.000	19:25	
Diagnostics Check	OK	19:25	

---

Data Download Date/Time: 04/04/2018 13:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WINTERS,ERIN,M  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: EDDINS,BRETT  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 04/04/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 640976  
DRY GAS EXPIRY DATE: 11/21/2019  
VIOLATION CODE: DUI  
ARREST TIME: 16:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: DUPONT

DATE/TIME OF TEST: 04/06/2018 18:18:34

---

Diagnostics Check	OK	18:10
Air Blank	0.000	18:11
Control Test	0.079	18:11
Air Blank	0.000	18:11
Subject Sample # 1	0.186	18:13
Breath Volume	1.445	18:13
Air Blank	0.000	18:14
Air Blank	0.000	18:16
Subject Sample # 2	0.183	18:16
Breath Volume	1.136	18:16
Air Blank	0.000	18:17
Control Test	0.076	18:17
Air Blank	0.000	18:18
Diagnostics Check	OK	18:18

---

Data Download Date/Time: 05/02/2018 17:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HENNESSEY, BRANDON, L  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: WALLACE, RYAN  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 04/04/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 640976  
DRY GAS EXPIRY DATE: 11/21/2019  
VIOLATION CODE: DUI  
ARREST TIME: 20:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: WALLACE

DATE/TIME OF TEST: 04/07/2018 20:53:38

---

Diagnostics Check	OK	20:50	
Air Blank	0.000	20:51	
Control Test	0.080	20:51	
Air Blank	0.000	20:51	
Subject Sample	# 1	20:52	Subject Test Refused
Breath Volume	0.000	20:52	
Air Blank	0.000	20:52	
Control Test	0.080	20:53	
Air Blank	0.000	20:53	
Diagnostics Check	OK	20:53	

---

Data Download Date/Time: 05/02/2018 17:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HENNESSEY, BRANDON, L  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: FLOYD, JOHN  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 04/04/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 640976  
DRY GAS EXPIRY DATE: 11/21/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: STAFFORD

DATE/TIME OF TEST: 04/12/2018 02:34:25

---

Diagnostics Check	OK	02:28
Air Blank	0.000	02:28
Control Test	0.080	02:28
Air Blank	0.000	02:29
Subject Sample # 1	0.248	02:29
Breath Volume	2.074	02:29
Air Blank	0.000	02:30
Air Blank	0.000	02:32
Subject Sample # 2	0.248	02:32
Breath Volume	2.265	02:32
Air Blank	0.000	02:33
Control Test	0.078	02:33
Air Blank	0.000	02:34
Diagnostics Check	OK	02:34

---

Data Download Date/Time: 05/02/2018 17:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HENNESSEY, BRANDON, L  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: HUSSEY, PATRICK  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 04/04/2018

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 640976  
DRY GAS EXPIRY DATE: 11/21/2019  
VIOLATION CODE: DUI  
ARREST TIME: 19:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:51

SUBJECT'S LAST NAME: VANDERBENT

DATE/TIME OF TEST: 04/16/2018 19:17:35

---

Diagnostics Check	OK	19:11
Air Blank	0.000	19:11
Control Test	0.080	19:12
Air Blank	0.000	19:12
Subject Sample # 1	0.225	19:13
Breath Volume	3.144	19:13
Air Blank	0.000	19:13
Air Blank	0.000	19:15
Subject Sample # 2	0.216	19:15
Breath Volume	3.023	19:15
Air Blank	0.000	19:16
Control Test	0.078	19:16
Air Blank	0.000	19:17
Diagnostics Check	OK	19:17

---

Data Download Date/Time: 05/02/2018 17:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAND,BRANDON,K  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: HAND,BRANDON  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 04/04/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 640976  
DRY GAS EXPIRY DATE: 11/21/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: GASPAR

DATE/TIME OF TEST: 04/19/2018 02:31:16

---

Diagnostics Check	OK	02:28	
Air Blank	0.000	02:28	
Control Test	0.081	02:29	
Air Blank	0.000	02:29	
Subject Sample	# 1	02:29	Subject Test Refused
Breath Volume	0.000	02:29	
Air Blank	0.000	02:30	
Control Test	0.080	02:30	
Air Blank	0.000	02:31	
Diagnostics Check	OK	02:31	

---

Data Download Date/Time: 05/02/2018 17:17



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CREPPEL,SEAN,M  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: CREPPEL,SEAN  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 04/04/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 640976  
DRY GAS EXPIRY DATE: 11/21/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: CONWAY

DATE/TIME OF TEST: 04/21/2018 03:25:38

---

Diagnostics Check	OK	03:22	
Air Blank	0.000	03:22	
Control Test	0.080	03:23	
Air Blank	0.000	03:23	
Subject Sample	# 1	03:24	Subject Test Refused
Breath Volume	0.000	03:24	
Air Blank	0.000	03:24	
Control Test	0.080	03:25	
Air Blank	0.000	03:25	
Diagnostics Check	OK	03:25	

---

Data Download Date/Time: 05/02/2018 17:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RADO,RYAN,R  
BREATH TEST OPERATOR AGENCY: St. Augustine PD  
ARRESTING OFFICER: RADO,RYAN  
ARRESTING OFFICER AGENCY: St. Augustine PD  
LAST AGENCY INSPECTION: 04/04/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 640976  
DRY GAS EXPIRY DATE: 11/21/2019  
VIOLATION CODE: DUI  
ARREST TIME: 18:04  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: ESTEVEZ

DATE/TIME OF TEST: 04/24/2018 18:54:15

---

Diagnostics Check	OK	18:42	
Air Blank	0.000	18:43	
Control Test	0.079	18:43	
Air Blank	0.000	18:43	
Subject Sample	# 1 0.125	18:47	Volume Not Met
Breath Volume	0.046	18:47	
Air Blank	0.000	18:47	
Air Blank	0.000	18:49	
Subject Sample	# 2 0.103	18:52	Volume Not Met
Breath Volume	0.476	18:52	
Air Blank	0.000	18:53	
Control Test	0.079	18:53	
Air Blank	0.000	18:54	
Diagnostics Check	OK	18:54	

---

Data Download Date/Time: 05/02/2018 17:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RADO,RYAN,R  
BREATH TEST OPERATOR AGENCY: St. Augustine PD  
ARRESTING OFFICER: RADO,RYAN  
ARRESTING OFFICER AGENCY: St. Augustine PD  
LAST AGENCY INSPECTION: 04/04/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 640976  
DRY GAS EXPIRY DATE: 11/21/2019  
VIOLATION CODE: DUI  
ARREST TIME: 18:04  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: ESTEVEZ

DATE/TIME OF TEST: 04/24/2018 19:10:45

---

Diagnostics Check		OK	18:56	
Air Blank		0.000	18:56	
Control Test		0.078	18:56	
Air Blank		0.000	18:57	
Subject Sample	# 1	0.111	19:00	Volume Not Met
Breath Volume		0.792	19:00	
Air Blank		0.000	19:01	
Air Blank		0.000	19:03	
Subject Sample	# 2	0.106	19:05	
Breath Volume		1.558	19:05	
Air Blank		0.000	19:06	
Air Blank		0.000	19:08	
Subject Sample	# 3	0.108	19:09	
Breath Volume		1.324	19:09	
Air Blank		0.000	19:09	
Control Test		0.077	19:10	
Air Blank		0.000	19:10	
Diagnostics Check		OK	19:10	

---

Data Download Date/Time: 05/02/2018 17:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HENNESSEY,BRANDON,L  
BREATH TEST OPERATOR AGENCY: St. Augustine Beach  
ARRESTING OFFICER: CORREA,RAFAEL  
ARRESTING OFFICER AGENCY: St. Augustine Beach  
LAST AGENCY INSPECTION: 04/04/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 640976  
DRY GAS EXPIRY DATE: 11/21/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 04/26/2018 23:12:12

---

Diagnostics Check	OK	23:05
Air Blank	0.000	23:06
Control Test	0.080	23:06
Air Blank	0.000	23:07
Subject Sample # 1	0.126	23:07
Breath Volume	2.039	23:07
Air Blank	0.000	23:08
Air Blank	0.000	23:10
Subject Sample # 2	0.115	23:10
Breath Volume	1.593	23:10
Air Blank	0.000	23:11
Control Test	0.078	23:11
Air Blank	0.000	23:12
Diagnostics Check	OK	23:12

---

Data Download Date/Time: 05/02/2018 17:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER,SCOTT,B  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: MILLER,SCOTT  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 04/04/2018

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 640976  
DRY GAS EXPIRY DATE: 11/21/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: DUNN

DATE/TIME OF TEST: 04/28/2018 00:40:03

---

Diagnostics Check	OK	00:37	
Air Blank	0.000	00:38	
Control Test	0.080	00:38	
Air Blank	0.000	00:39	
Subject Sample	# 1	00:39	RFI Detect
Breath Volume	0.000	00:39	
Air Blank		00:39	RFI Detect
Air Blank	0.000	00:40	

---

Data Download Date/Time: 05/02/2018 17:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEDINA,MARMILLY,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: LOUGHRAN,KEVIN  
ARRESTING OFFICER AGENCY: OTHER Local  
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 562838  
DRY GAS EXPIRY DATE: 09/06/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 04/02/2018 04:12:51

---

Diagnostics Check	OK	04:05
Air Blank	0.000	04:06
Control Test	0.082	04:06
Air Blank	0.000	04:07
Subject Sample # 1	0.193	04:08
Breath Volume	2.015	04:08
Air Blank	0.000	04:08
Air Blank	0.000	04:10
Subject Sample # 2	0.183	04:11
Breath Volume	1.792	04:11
Air Blank	0.000	04:11
Control Test	0.081	04:12
Air Blank	0.000	04:12
Diagnostics Check	OK	04:12

---

Data Download Date/Time: 04/26/2018 12:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEDINA,MARMILLY,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: SACKMANN,ANDREW  
ARRESTING OFFICER AGENCY: OTHER State  
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 562838  
DRY GAS EXPIRY DATE: 09/06/2019  
VIOLATION CODE: DUI  
ARREST TIME: 20:42  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 04/02/2018 22:09:42

---

Diagnostics Check	OK	22:04	
Air Blank	0.000	22:04	
Control Test	0.079	22:05	
Air Blank	0.000	22:05	
Subject Sample	# 1	22:08	Subject Test Refused
Breath Volume	0.000	22:08	
Air Blank	0.000	22:08	
Control Test	0.079	22:09	
Air Blank	0.000	22:09	
Diagnostics Check	OK	22:09	

---

Data Download Date/Time: 04/26/2018 12:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: FITZPATRICK,LACEY  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 562838  
DRY GAS EXPIRY DATE: 09/06/2019  
VIOLATION CODE: DUI  
ARREST TIME: 10:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:33

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 04/03/2018 12:00:10

---

Diagnostics Check	OK	11:54
Air Blank	0.000	11:54
Control Test	0.083	11:54
Air Blank	0.000	11:55
Subject Sample # 1	0.000	11:55
Breath Volume	1.160	11:55
Air Blank	0.000	11:56
Air Blank	0.000	11:58
Subject Sample # 2	0.000	11:58
Breath Volume	1.261	11:58
Air Blank	0.000	11:59
Control Test	0.082	11:59
Air Blank	0.000	12:00
Diagnostics Check	OK	12:00

---

Data Download Date/Time: 04/26/2018 12:17



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEDINA,MARMILLY,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: BOHM,JONATHAN  
ARRESTING OFFICER AGENCY: Fort Lauderdale PD  
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 562838  
DRY GAS EXPIRY DATE: 09/06/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: VASQUEZ

DATE/TIME OF TEST: 04/07/2018 03:23:13

---

Diagnostics Check	OK	03:17
Air Blank	0.000	03:17
Control Test	0.080	03:17
Air Blank	0.000	03:18
Subject Sample # 1	0.165	03:18
Breath Volume	1.656	03:18
Air Blank	0.000	03:19
Air Blank	0.000	03:21
Subject Sample # 2	0.169	03:21
Breath Volume	1.710	03:21
Air Blank	0.000	03:22
Control Test	0.080	03:22
Air Blank	0.000	03:23
Diagnostics Check	OK	03:23

---

Data Download Date/Time: 04/26/2018 12:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEDINA,MARMILLY,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: KEATING,KAYLA  
ARRESTING OFFICER AGENCY: Fort Lauderdale PD  
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 562838  
DRY GAS EXPIRY DATE: 09/06/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: LOWE

DATE/TIME OF TEST: 04/07/2018 23:41:46

---

Diagnostics Check	OK	23:35
Air Blank	0.000	23:35
Control Test	0.082	23:36
Air Blank	0.000	23:36
Subject Sample # 1	0.164	23:37
Breath Volume	1.664	23:37
Air Blank	0.000	23:37
Air Blank	0.000	23:39
Subject Sample # 2	0.167	23:40
Breath Volume	1.472	23:40
Air Blank	0.000	23:40
Control Test	0.082	23:41
Air Blank	0.000	23:41
Diagnostics Check	OK	23:41

---

Data Download Date/Time: 04/26/2018 12:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAZIER,ANAYA,S  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: BUDISH,HARRISON  
ARRESTING OFFICER AGENCY: Pembroke Pines PD  
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 562838  
DRY GAS EXPIRY DATE: 09/06/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 04/10/2018 04:43:02

---

Diagnostics Check	OK	04:36
Air Blank	0.000	04:36
Control Test	0.082	04:36
Air Blank	0.000	04:37
Subject Sample # 1	0.184	04:38
Breath Volume	1.816	04:38
Air Blank	0.000	04:38
Air Blank	0.000	04:40
Subject Sample # 2	0.178	04:41
Breath Volume	2.066	04:41
Air Blank	0.000	04:42
Control Test	0.082	04:42
Air Blank	0.000	04:42
Diagnostics Check	OK	04:43

---

Data Download Date/Time: 04/26/2018 12:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TILLMAN,SHEILA,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: TRABUE,TIMOTHY  
ARRESTING OFFICER AGENCY: Pembroke Pines PD  
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 562838  
DRY GAS EXPIRY DATE: 09/06/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: VANCE JR

DATE/TIME OF TEST: 04/11/2018 04:32:16

---

Diagnostics Check		OK	04:18	
Air Blank		0.000	04:18	
Control Test		0.082	04:19	
Air Blank		0.000	04:19	
Subject Sample	# 1	0.172	04:21	
Breath Volume		1.617	04:21	
Air Blank		0.000	04:21	
Air Blank		0.000	04:23	
Subject Sample	# 2	0.133	04:27	Volume Not Met
Breath Volume		0.398	04:27	
Air Blank		0.000	04:27	
Air Blank		0.000	04:29	
Subject Sample	# 3	0.162	04:30	
Breath Volume		1.812	04:30	
Air Blank		0.000	04:31	
Control Test		0.081	04:31	
Air Blank		0.000	04:32	
Diagnostics Check		OK	04:32	

---

Data Download Date/Time: 04/26/2018 12:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TILLMAN,SHEILA,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: RAZI,UMAIR  
ARRESTING OFFICER AGENCY: Pembroke Pines PD  
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 562838  
DRY GAS EXPIRY DATE: 09/06/2019  
VIOLATION CODE: DUI  
ARREST TIME: 04:10  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:07

SUBJECT'S LAST NAME: SUCKIE

DATE/TIME OF TEST: 04/11/2018 06:29:28

---

Diagnostics Check	OK	06:22
Air Blank	0.000	06:22
Control Test	0.082	06:22
Air Blank	0.000	06:23
Subject Sample # 1	0.157	06:24
Breath Volume	1.152	06:24
Air Blank	0.000	06:25
Air Blank	0.000	06:27
Subject Sample # 2	0.163	06:28
Breath Volume	1.246	06:28
Air Blank	0.000	06:28
Control Test	0.081	06:28
Air Blank	0.000	06:29
Diagnostics Check	OK	06:29

---

Data Download Date/Time: 04/26/2018 12:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TILLMAN,SHEILA,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: CARTER,JAMES  
ARRESTING OFFICER AGENCY: Fort Lauderdale PD  
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 562838  
DRY GAS EXPIRY DATE: 09/06/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 04/13/2018 02:00:52

---

Diagnostics Check	OK	01:51
Air Blank	0.000	01:52
Control Test	0.082	01:52
Air Blank	0.000	01:52
Subject Sample # 1	0.128	01:54
Breath Volume	1.355	01:54
Air Blank	0.000	01:55
Air Blank	0.000	01:57
Subject Sample # 2	0.123	01:59
Breath Volume	1.625	01:59
Air Blank	0.000	01:59
Control Test	0.082	02:00
Air Blank	0.000	02:00
Diagnostics Check	OK	02:00

---

Data Download Date/Time: 04/26/2018 12:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TILLMAN,SHEILA,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: HENNESSEY,JONATHAN  
ARRESTING OFFICER AGENCY: Sunrise PD  
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 562838  
DRY GAS EXPIRY DATE: 09/06/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: JIMINEZ

DATE/TIME OF TEST: 04/14/2018 05:10:59

---

Diagnostics Check	OK	05:02
Air Blank	0.000	05:03
Control Test	0.082	05:03
Air Blank	0.000	05:04
Subject Sample # 1	0.129	05:06
Breath Volume	2.300	05:06
Air Blank	0.000	05:07
Air Blank	0.000	05:09
Subject Sample # 2	0.131	05:09
Breath Volume	2.378	05:09
Air Blank	0.000	05:10
Control Test	0.081	05:10
Air Blank	0.000	05:10
Diagnostics Check	OK	05:10

---

Data Download Date/Time: 04/26/2018 12:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TILLMAN,SHEILA,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: O'DOR,ROBERT  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 562838  
DRY GAS EXPIRY DATE: 09/06/2019  
VIOLATION CODE: DUI  
ARREST TIME: 05:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:30

SUBJECT'S LAST NAME: SALERO JR

DATE/TIME OF TEST: 04/15/2018 08:56:57

---

Diagnostics Check	OK	08:50
Air Blank	0.000	08:51
Control Test	0.081	08:51
Air Blank	0.000	08:51
Subject Sample # 1	0.144	08:52
Breath Volume	2.875	08:52
Air Blank	0.000	08:52
Air Blank	0.000	08:54
Subject Sample # 2	0.139	08:55
Breath Volume	1.312	08:55
Air Blank	0.000	08:56
Control Test	0.081	08:56
Air Blank	0.000	08:56
Diagnostics Check	OK	08:56

---

Data Download Date/Time: 04/26/2018 12:17



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAZIER,ANAYA,S  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: HARRIS,MILTON  
ARRESTING OFFICER AGENCY: Pembroke Pines PD  
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 562838  
DRY GAS EXPIRY DATE: 09/06/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 04/16/2018 02:15:34

---

Diagnostics Check	OK	02:09
Air Blank	0.000	02:09
Control Test	0.081	02:09
Air Blank	0.000	02:10
Subject Sample # 1	0.216	02:11
Breath Volume	1.460	02:11
Air Blank	0.000	02:11
Air Blank	0.000	02:13
Subject Sample # 2	0.216	02:14
Breath Volume	2.304	02:14
Air Blank	0.000	02:14
Control Test	0.081	02:14
Air Blank	0.000	02:15
Diagnostics Check	OK	02:15

---

Data Download Date/Time: 04/26/2018 12:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRAY,SANDRA,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: KUTIKOFF,ADAM  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 562838  
DRY GAS EXPIRY DATE: 09/06/2019  
VIOLATION CODE: DUI  
ARREST TIME: 13:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:24

SUBJECT'S LAST NAME: LAMARRE

DATE/TIME OF TEST: 04/18/2018 16:25:16

---

Diagnostics Check	OK	16:18
Air Blank	0.000	16:19
Control Test	0.081	16:19
Air Blank	0.000	16:19
Subject Sample # 1	0.201	16:20
Breath Volume	1.531	16:20
Air Blank	0.000	16:21
Air Blank	0.000	16:23
Subject Sample # 2	0.205	16:23
Breath Volume	1.343	16:23
Air Blank	0.000	16:24
Control Test	0.082	16:24
Air Blank	0.000	16:25
Diagnostics Check	OK	16:25

---

Data Download Date/Time: 04/26/2018 12:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TILLMAN,SHEILA,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: CAMPBELL,DWAYNE  
ARRESTING OFFICER AGENCY: Miramar PD  
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 562838  
DRY GAS EXPIRY DATE: 09/06/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:28  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: VARGAS ALFARO

DATE/TIME OF TEST: 04/20/2018 01:48:35

---

Diagnostics Check	OK	01:41
Air Blank	0.000	01:42
Control Test	0.081	01:42
Air Blank	0.000	01:42
Subject Sample # 1	0.142	01:44
Breath Volume	1.425	01:44
Air Blank	0.000	01:44
Air Blank	0.000	01:46
Subject Sample # 2	0.123	01:47
Breath Volume	1.511	01:47
Air Blank	0.000	01:47
Control Test	0.080	01:48
Air Blank	0.000	01:48
Diagnostics Check	OK	01:48

---

Data Download Date/Time: 04/26/2018 12:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRAY,SANDRA,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: ROJAS,RANDY  
ARRESTING OFFICER AGENCY: OTHER State  
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 562838  
DRY GAS EXPIRY DATE: 09/06/2019  
VIOLATION CODE: DUI  
ARREST TIME: 04:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:57

SUBJECT'S LAST NAME: PARRISH

DATE/TIME OF TEST: 04/20/2018 05:23:30

---

Diagnostics Check	OK	05:17
Air Blank	0.000	05:17
Control Test	0.081	05:18
Air Blank	0.000	05:18
Subject Sample # 1	0.086	05:19
Breath Volume	3.746	05:19
Air Blank	0.000	05:19
Air Blank	0.000	05:21
Subject Sample # 2	0.085	05:22
Breath Volume	3.886	05:22
Air Blank	0.000	05:22
Control Test	0.080	05:22
Air Blank	0.000	05:23
Diagnostics Check	OK	05:23

---

Data Download Date/Time: 04/26/2018 12:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TILLMAN,SHEILA,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: FLINK,ANDREW  
ARRESTING OFFICER AGENCY: Miramar PD  
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 562838  
DRY GAS EXPIRY DATE: 09/06/2019  
VIOLATION CODE: DUI  
ARREST TIME: 04:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: ALTES

DATE/TIME OF TEST: 04/21/2018 05:50:18

---

Diagnostics Check	OK	05:43
Air Blank	0.000	05:44
Control Test	0.082	05:44
Air Blank	0.000	05:44
Subject Sample # 1	0.161	05:45
Breath Volume	1.160	05:45
Air Blank	0.000	05:46
Air Blank	0.000	05:48
Subject Sample # 2	0.160	05:48
Breath Volume	1.285	05:48
Air Blank	0.000	05:49
Control Test	0.081	05:49
Air Blank	0.000	05:50
Diagnostics Check	OK	05:50

---

Data Download Date/Time: 04/26/2018 12:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEDINA,MARMILLY,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: ORDAZ,ALEXANDER  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 562838  
DRY GAS EXPIRY DATE: 09/06/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:27  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: EBANKS

DATE/TIME OF TEST: 04/23/2018 03:22:08

---

Diagnostics Check		OK	03:15
Air Blank		0.000	03:16
Control Test		0.082	03:16
Air Blank		0.000	03:17
Subject Sample	# 1	0.205	03:17
Breath Volume		1.921	03:17
Air Blank		0.000	03:18
Air Blank		0.000	03:20
Subject Sample	# 2	0.191	03:20
Breath Volume		1.531	03:20
Air Blank		0.000	03:21
Control Test		0.082	03:21
Air Blank		0.000	03:22
Diagnostics Check		OK	03:22

---

Data Download Date/Time: 04/26/2018 12:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAZIER,ANAYA,S  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: MALDONADO,ANGEL  
ARRESTING OFFICER AGENCY: Margate PD  
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 562838  
DRY GAS EXPIRY DATE: 09/06/2019  
VIOLATION CODE: DUI  
ARREST TIME: 20:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: CHANG

DATE/TIME OF TEST: 04/23/2018 22:10:50

---

Diagnostics Check	OK	22:02
Air Blank	0.000	22:02
Control Test	0.081	22:02
Air Blank	0.000	22:03
Subject Sample # 1	0.000	22:04
Breath Volume	1.121	22:04
Air Blank	0.000	22:04
Air Blank	0.000	22:06
Subject Sample # 2	0.000	22:09
Breath Volume	1.156	22:09
Air Blank	0.000	22:09
Control Test	0.082	22:10
Air Blank	0.000	22:10
Diagnostics Check	OK	22:10

---

Data Download Date/Time: 04/26/2018 12:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRAY,SANDRA,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: BRETON,JESSIKA  
ARRESTING OFFICER AGENCY: Pembroke Pines PD  
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 562838  
DRY GAS EXPIRY DATE: 09/06/2019  
VIOLATION CODE: DUI  
ARREST TIME: 04:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: CASTRO

DATE/TIME OF TEST: 04/26/2018 05:26:27

---

Diagnostics Check		OK	05:16
Air Blank		0.000	05:16
Control Test		0.082	05:17
Air Blank		0.000	05:17
Subject Sample	# 1	0.178	05:18
Breath Volume		1.433	05:18
Air Blank		0.000	05:18
Air Blank		0.000	05:20
Subject Sample	# 2	0.201	05:21
Breath Volume		3.503	05:21
Air Blank		0.000	05:22
Air Blank		0.000	05:24
Subject Sample	# 3	0.187	05:24
Breath Volume		1.429	05:24
Air Blank		0.000	05:25
Control Test		0.081	05:25
Air Blank		0.000	05:26
Diagnostics Check		OK	05:26

---

Data Download Date/Time: 04/26/2018 12:17



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TILLMAN,SHEILA,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: BOONE,MICHEAL  
ARRESTING OFFICER AGENCY: Pembroke Pines PD  
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 562838  
DRY GAS EXPIRY DATE: 09/06/2019  
VIOLATION CODE: DUI  
ARREST TIME: 04:21  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: BAEZ

DATE/TIME OF TEST: 04/26/2018 05:50:24

---

Diagnostics Check		OK	05:35	
Air Blank		0.000	05:36	
Control Test		0.081	05:36	
Air Blank		0.000	05:37	
Subject Sample	# 1	0.176	05:40	Volume Not Met
Breath Volume		0.167	05:40	
Air Blank		0.000	05:40	
Air Blank		0.000	05:42	
Subject Sample	# 2	0.210	05:43	
Breath Volume		1.496	05:43	
Air Blank		0.000	05:43	
Air Blank		0.000	05:45	
Subject Sample	# 3	0.187	05:48	Volume Not Met
Breath Volume		0.960	05:48	
Air Blank		0.000	05:49	
Control Test		0.081	05:49	
Air Blank		0.000	05:50	
Diagnostics Check		OK	05:50	

---

Data Download Date/Time: 04/26/2018 12:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TILLMAN,SHEILA,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: BOONE,MICHAEL  
ARRESTING OFFICER AGENCY: Pembroke Pines PD  
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 562838  
DRY GAS EXPIRY DATE: 09/06/2019  
VIOLATION CODE: DUI  
ARREST TIME: 04:21  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: BAEZ

DATE/TIME OF TEST: 04/26/2018 06:00:56

---

Diagnostics Check	OK	05:52
Air Blank	0.000	05:52
Control Test	0.081	05:52
Air Blank	0.000	05:53
Subject Sample # 1	0.202	05:54
Breath Volume	1.406	05:54
Air Blank	0.000	05:54
Air Blank	0.000	05:56
Subject Sample # 2	0.200	05:59
Breath Volume	1.957	05:59
Air Blank	0.000	06:00
Control Test	0.081	06:00
Air Blank	0.000	06:00
Diagnostics Check	OK	06:00

---

Data Download Date/Time: 04/26/2018 12:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRAY, SANDRA,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: LOUIS, WALEX  
ARRESTING OFFICER AGENCY: OTHER State  
LAST AGENCY INSPECTION: 04/26/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 562838  
DRY GAS EXPIRY DATE: 09/06/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: GOLDWIRE

DATE/TIME OF TEST: 04/27/2018 02:12:05

---

Diagnostics Check	OK	02:05
Air Blank	0.000	02:06
Control Test	0.083	02:06
Air Blank	0.000	02:07
Subject Sample # 1	0.256	02:07
Breath Volume	1.890	02:07
Air Blank	0.000	02:08
Air Blank	0.000	02:10
Subject Sample # 2	0.251	02:10
Breath Volume	2.542	02:10
Air Blank	0.000	02:11
Control Test	0.082	02:11
Air Blank	0.000	02:12
Diagnostics Check	OK	02:12

---

Data Download Date/Time: 05/10/2018 12:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TILLMAN,SHEILA,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: SAPP,JOSHUA  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 04/26/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 562838  
DRY GAS EXPIRY DATE: 09/06/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: DOUGLAS

DATE/TIME OF TEST: 04/27/2018 03:43:14

---

Diagnostics Check	OK	03:30
Air Blank	0.000	03:31
Control Test	0.082	03:31
Air Blank	0.000	03:31
Subject Sample # 1	0.264	03:33
Breath Volume	1.589	03:33
Air Blank	0.000	03:34
Air Blank	0.000	03:35
Subject Sample # 2	0.241	03:38
Breath Volume	1.468	03:38
Air Blank	0.000	03:38
Air Blank	0.000	03:40
Subject Sample # 3	0.240	03:41
Breath Volume	1.691	03:41
Air Blank	0.000	03:42
Control Test	0.081	03:42
Air Blank	0.000	03:43
Diagnostics Check	OK	03:43

---

Data Download Date/Time: 05/10/2018 12:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEDINA,MARMILLY,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: JOSEPH,WILLIAMSON  
ARRESTING OFFICER AGENCY: OTHER State  
LAST AGENCY INSPECTION: 04/26/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 562838  
DRY GAS EXPIRY DATE: 09/06/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: LECHE

DATE/TIME OF TEST: 04/28/2018 03:52:29

---

Diagnostics Check	OK	03:45
Air Blank	0.000	03:46
Control Test	0.083	03:46
Air Blank	0.000	03:47
Subject Sample # 1	0.200	03:47
Breath Volume	2.062	03:47
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 2	0.202	03:51
Breath Volume	2.269	03:51
Air Blank	0.000	03:51
Control Test	0.080	03:51
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

---

Data Download Date/Time: 05/10/2018 12:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEDINA,MARMILLY,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: BOHM,JONATHAN  
ARRESTING OFFICER AGENCY: Fort Lauderdale PD  
LAST AGENCY INSPECTION: 04/26/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 562838  
DRY GAS EXPIRY DATE: 09/06/2019  
VIOLATION CODE: DUI  
ARREST TIME: 20:58  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: GUTIERREZ ROBLERO

DATE/TIME OF TEST: 04/28/2018 22:29:28

---

Diagnostics Check	OK	22:21
Air Blank	0.000	22:21
Control Test	0.082	22:21
Air Blank	0.000	22:22
Subject Sample # 1	0.227	22:23
Breath Volume	1.238	22:23
Air Blank	0.000	22:24
Air Blank	0.000	22:26
Subject Sample # 2	0.237	22:27
Breath Volume	1.605	22:27
Air Blank	0.000	22:28
Control Test	0.081	22:28
Air Blank	0.000	22:29
Diagnostics Check	OK	22:29

---

Data Download Date/Time: 05/10/2018 12:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEDINA,MARMILLY,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: HENNESSY,JONATHAN  
ARRESTING OFFICER AGENCY: Sunrise PD  
LAST AGENCY INSPECTION: 04/26/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 562838  
DRY GAS EXPIRY DATE: 09/06/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: STACK

DATE/TIME OF TEST: 04/29/2018 03:33:22

---

Diagnostics Check	OK	03:26
Air Blank	0.000	03:27
Control Test	0.081	03:27
Air Blank	0.000	03:28
Subject Sample # 1	0.205	03:28
Breath Volume	1.875	03:28
Air Blank	0.000	03:29
Air Blank	0.000	03:31
Subject Sample # 2	0.186	03:31
Breath Volume	1.554	03:31
Air Blank	0.000	03:32
Control Test	0.081	03:32
Air Blank	0.000	03:33
Diagnostics Check	OK	03:33

---

Data Download Date/Time: 05/10/2018 12:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEDINA,MARMILLY,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: CARTER,JAMES  
ARRESTING OFFICER AGENCY: Fort Lauderdale PD  
LAST AGENCY INSPECTION: 04/26/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 562838  
DRY GAS EXPIRY DATE: 09/06/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: PINEDA GRADIZ

DATE/TIME OF TEST: 04/29/2018 04:45:32

---

Diagnostics Check	OK	04:39
Air Blank	0.000	04:39
Control Test	0.081	04:40
Air Blank	0.000	04:40
Subject Sample # 1	0.187	04:41
Breath Volume	2.332	04:41
Air Blank	0.000	04:41
Air Blank	0.000	04:43
Subject Sample # 2	0.186	04:44
Breath Volume	2.210	04:44
Air Blank	0.000	04:44
Control Test	0.081	04:44
Air Blank	0.000	04:45
Diagnostics Check	OK	04:45

---

Data Download Date/Time: 05/10/2018 12:08



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TILLMAN,SHEILA,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: ORTIZ,JEFFREY  
ARRESTING OFFICER AGENCY: Hollywood PD  
LAST AGENCY INSPECTION: 04/26/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 562838  
DRY GAS EXPIRY DATE: 09/06/2019  
VIOLATION CODE: DUI  
ARREST TIME: 08:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:25

SUBJECT'S LAST NAME: JARQUIN OBANDO

DATE/TIME OF TEST: 04/29/2018 11:17:05

---

Diagnostics Check	OK	11:09
Air Blank	0.000	11:10
Control Test	0.082	11:10
Air Blank	0.000	11:11
Subject Sample # 1	0.262	11:12
Breath Volume	1.484	11:12
Air Blank	0.000	11:12
Air Blank	0.000	11:14
Subject Sample # 2	0.255	11:15
Breath Volume	2.261	11:15
Air Blank	0.000	11:16
Control Test	0.081	11:16
Air Blank	0.000	11:17
Diagnostics Check	OK	11:17

---

Data Download Date/Time: 05/10/2018 12:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEDINA,MARMILLY,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: DEWEESE,MATTHEW  
ARRESTING OFFICER AGENCY: Sunrise PD  
LAST AGENCY INSPECTION: 04/26/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 562838  
DRY GAS EXPIRY DATE: 09/06/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: ARANA

DATE/TIME OF TEST: 04/30/2018 00:46:51

---

Diagnostics Check	OK	00:40
Air Blank	0.000	00:40
Control Test	0.082	00:41
Air Blank	0.000	00:41
Subject Sample # 1	0.204	00:42
Breath Volume	1.625	00:42
Air Blank	0.000	00:42
Air Blank	0.000	00:44
Subject Sample # 2	0.194	00:45
Breath Volume	1.371	00:45
Air Blank	0.000	00:45
Control Test	0.081	00:46
Air Blank	0.000	00:46
Diagnostics Check	OK	00:46

---

Data Download Date/Time: 05/10/2018 12:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TUCKER, WILLIAM,  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: TUCKER, WILLIAM  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 03/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 390314  
DRY GAS EXPIRY DATE: 04/19/2019  
VIOLATION CODE: DUI  
ARREST TIME: 20:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:18

SUBJECT'S LAST NAME: KIOKO

DATE/TIME OF TEST: 04/08/2018 20:50:01

---

Diagnostics Check	OK	20:43
Air Blank	0.000	20:44
Control Test	0.081	20:44
Air Blank	0.000	20:45
Subject Sample # 1	0.194	20:45
Breath Volume	3.058	20:45
Air Blank	0.000	20:46
Air Blank	0.000	20:47
Subject Sample # 2	0.192	20:48
Breath Volume	2.703	20:48
Air Blank	0.000	20:49
Control Test	0.081	20:49
Air Blank	0.000	20:49
Diagnostics Check	OK	20:50

---

Data Download Date/Time: 04/15/2018 17:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001152

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PAGE,BRIAN,R  
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD  
ARRESTING OFFICER: JENKINS,ANDRE  
ARRESTING OFFICER AGENCY: New Smyrna Beach PD  
LAST AGENCY INSPECTION: 03/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A2  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: ANTHONY

DATE/TIME OF TEST: 04/07/2018 02:45:00

---

Diagnostics Check	OK	02:38
Air Blank	0.000	02:39
Control Test	0.079	02:39
Air Blank	0.000	02:39
Subject Sample # 1	0.261	02:40
Breath Volume	2.722	02:40
Air Blank	0.000	02:41
Air Blank	0.000	02:42
Subject Sample # 2	0.257	02:43
Breath Volume	2.976	02:43
Air Blank	0.000	02:44
Control Test	0.078	02:44
Air Blank	0.000	02:44
Diagnostics Check	OK	02:45

---

Data Download Date/Time: 04/29/2018 09:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001152

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PETERSON,OFFICER,  
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD  
ARRESTING OFFICER: MALDONADO,OFFICER  
ARRESTING OFFICER AGENCY: New Smyrna Beach PD  
LAST AGENCY INSPECTION: 03/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A2  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 18:03  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: GICK

DATE/TIME OF TEST: 04/09/2018 18:43:19

---

Diagnostics Check	OK	18:40	
Air Blank	0.000	18:40	
Control Test	0.080	18:41	
Air Blank	0.000	18:41	
Subject Sample	# 1	18:41	Subject Test Refused
Breath Volume	0.000	18:41	
Air Blank	0.000	18:42	
Control Test	0.079	18:42	
Air Blank	0.000	18:43	
Diagnostics Check	OK	18:43	

---

Data Download Date/Time: 04/29/2018 09:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001152

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PETERSON,OFFICER,  
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD  
ARRESTING OFFICER: WASHINGTON,OFFICER  
ARRESTING OFFICER AGENCY: New Smyrna Beach PD  
LAST AGENCY INSPECTION: 03/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A2  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 17:35  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: CORTES

DATE/TIME OF TEST: 04/13/2018 18:37:14

---

Diagnostics Check	OK	18:21	
Air Blank	0.000	18:21	
Control Test	0.079	18:22	
Air Blank	0.000	18:22	
Subject Sample # 1	0.275	18:25	
Breath Volume	1.363	18:25	
Air Blank	0.000	18:26	
Air Blank	0.000	18:27	
Subject Sample # 2		18:31	No Sample Provided
Breath Volume	0.000	18:31	
Air Blank	0.000	18:31	
Air Blank	0.000	18:33	
Subject Sample # 3		18:35	Subject Test Refused
Breath Volume	0.019	18:35	
Air Blank	0.000	18:35	
Control Test	0.079	18:36	
Air Blank	0.000	18:36	
Diagnostics Check	Fail	18:36	Voltage/Current Test
Air Blank	0.000	18:37	

---

Data Download Date/Time: 04/29/2018 09:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001152

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS,DAVID,B  
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD  
ARRESTING OFFICER: WILLIAMS,DAVID  
ARRESTING OFFICER AGENCY: New Smyrna Beach PD  
LAST AGENCY INSPECTION: 03/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A2  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 23:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: SPENCER

DATE/TIME OF TEST: 04/15/2018 00:23:04

---

Diagnostics Check	OK	00:15
Air Blank	0.000	00:16
Control Test	0.079	00:16
Air Blank	0.000	00:16
Subject Sample # 1	0.203	00:18
Breath Volume	1.785	00:18
Air Blank	0.000	00:19
Air Blank	0.000	00:20
Subject Sample # 2	0.202	00:21
Breath Volume	2.519	00:21
Air Blank	0.000	00:22
Control Test	0.078	00:22
Air Blank	0.000	00:22
Diagnostics Check	OK	00:23

---

Data Download Date/Time: 04/29/2018 09:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001152

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REVE,JASON,R  
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD  
ARRESTING OFFICER: CILENTI,DANIEL  
ARRESTING OFFICER AGENCY: New Smyrna Beach PD  
LAST AGENCY INSPECTION: 03/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A2  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 06:44  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 06:55

SUBJECT'S LAST NAME: LEVINE

DATE/TIME OF TEST: 04/20/2018 07:18:24

---

Diagnostics Check	OK	07:15	
Air Blank	0.000	07:16	
Control Test	0.079	07:16	
Air Blank	0.000	07:16	
Subject Sample	# 1	07:17	Subject Test Refused
Breath Volume	0.000	07:17	
Air Blank	0.000	07:17	
Control Test	0.079	07:17	
Air Blank	0.000	07:18	
Diagnostics Check	OK	07:18	

---

Data Download Date/Time: 04/29/2018 09:51



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RITTENOUR,COREY,T  
BREATH TEST OPERATOR AGENCY: Volusia County SO  
ARRESTING OFFICER: COLEY,NICOLEA  
ARRESTING OFFICER AGENCY: Holly Hills PD  
LAST AGENCY INSPECTION: 04/16/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 644476  
DRY GAS EXPIRY DATE: 12/01/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: DOOLEY

DATE/TIME OF TEST: 04/18/2018 02:25:28

---

Diagnostics Check	OK	02:21	
Air Blank	0.000	02:22	
Control Test	0.081	02:22	
Air Blank	0.000	02:23	
Subject Sample	# 1	02:24	Subject Test Refused
Breath Volume	0.000	02:24	
Air Blank	0.000	02:24	
Control Test	0.080	02:24	
Air Blank	0.000	02:25	
Diagnostics Check	OK	02:25	

---

Data Download Date/Time: 05/30/2018 11:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CORBIN,STEPHEN,A  
BREATH TEST OPERATOR AGENCY: Volusia County SO  
ARRESTING OFFICER: CORBIN,STEPHEN  
ARRESTING OFFICER AGENCY: Volusia County SO  
LAST AGENCY INSPECTION: 04/16/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 644476  
DRY GAS EXPIRY DATE: 12/01/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:05  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: MYERS

DATE/TIME OF TEST: 04/21/2018 03:02:53

---

Diagnostics Check	OK	02:55
Air Blank	0.000	02:56
Control Test	0.081	02:56
Air Blank	0.000	02:57
Subject Sample # 1	0.257	02:58
Breath Volume	2.574	02:58
Air Blank	0.000	02:59
Air Blank	0.000	03:00
Subject Sample # 2	0.251	03:01
Breath Volume	2.160	03:01
Air Blank	0.000	03:01
Control Test	0.081	03:02
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

---

Data Download Date/Time: 05/30/2018 11:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PECK,KEITH,D  
BREATH TEST OPERATOR AGENCY: Volusia County SO  
ARRESTING OFFICER: PECK,KEITH  
ARRESTING OFFICER AGENCY: Volusia County SO  
LAST AGENCY INSPECTION: 04/16/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 644476  
DRY GAS EXPIRY DATE: 12/01/2019  
VIOLATION CODE: DUI  
ARREST TIME: 19:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:44

SUBJECT'S LAST NAME: LEBARON

DATE/TIME OF TEST: 04/22/2018 20:42:56

---

Diagnostics Check	OK	20:40	
Air Blank	0.000	20:40	
Control Test	0.082	20:40	
Air Blank	0.000	20:41	
Subject Sample	# 1	20:41	Subject Test Refused
Breath Volume	0.000	20:41	
Air Blank	0.000	20:42	
Control Test	0.081	20:42	
Air Blank	0.000	20:42	
Diagnostics Check	OK	20:42	

---

Data Download Date/Time: 05/30/2018 11:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PECK,KEITH,D  
BREATH TEST OPERATOR AGENCY: Volusia County SO  
ARRESTING OFFICER: REINHARDT,THOMAS  
ARRESTING OFFICER AGENCY: Volusia County SO  
LAST AGENCY INSPECTION: 04/16/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 644476  
DRY GAS EXPIRY DATE: 12/01/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: HARPER

DATE/TIME OF TEST: 04/28/2018 04:28:08

---

Diagnostics Check	OK	04:20
Air Blank	0.000	04:20
Control Test	0.081	04:20
Air Blank	0.000	04:21
Subject Sample # 1	0.209	04:21
Breath Volume	1.425	04:21
Air Blank	0.000	04:22
Air Blank	0.000	04:24
Subject Sample # 2	0.227	04:26
Breath Volume	3.019	04:26
Air Blank	0.000	04:27
Control Test	0.079	04:27
Air Blank	0.000	04:28
Diagnostics Check	OK	04:28

---

Data Download Date/Time: 05/30/2018 11:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RUBIN,RICHARD,D  
BREATH TEST OPERATOR AGENCY: Ormond Beach PD  
ARRESTING OFFICER: RAIHALA,J  
ARRESTING OFFICER AGENCY: Ormond Beach PD  
LAST AGENCY INSPECTION: 04/16/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 644476  
DRY GAS EXPIRY DATE: 12/01/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:46

SUBJECT'S LAST NAME: TESTA

DATE/TIME OF TEST: 04/28/2018 23:12:56

---

Diagnostics Check	OK	23:06
Air Blank	0.000	23:07
Control Test	0.082	23:07
Air Blank	0.000	23:08
Subject Sample # 1	0.199	23:08
Breath Volume	3.136	23:08
Air Blank	0.000	23:09
Air Blank	0.000	23:10
Subject Sample # 2	0.201	23:11
Breath Volume	3.070	23:11
Air Blank	0.000	23:12
Control Test	0.080	23:12
Air Blank	0.000	23:12
Diagnostics Check	OK	23:12

---

Data Download Date/Time: 05/30/2018 11:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PECK,KEITH,D  
BREATH TEST OPERATOR AGENCY: Volusia County SO  
ARRESTING OFFICER: RAIHALA,JON  
ARRESTING OFFICER AGENCY: Ormond Beach PD  
LAST AGENCY INSPECTION: 04/16/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 644476  
DRY GAS EXPIRY DATE: 12/01/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: MAC KENZIE

DATE/TIME OF TEST: 04/30/2018 02:52:09

---

Diagnostics Check	OK	02:45
Air Blank	0.000	02:45
Control Test	0.082	02:46
Air Blank	0.000	02:46
Subject Sample # 1	0.223	02:47
Breath Volume	1.945	02:47
Air Blank	0.000	02:48
Air Blank	0.000	02:49
Subject Sample # 2	0.222	02:50
Breath Volume	2.917	02:50
Air Blank	0.000	02:51
Control Test	0.081	02:51
Air Blank	0.000	02:52
Diagnostics Check	OK	02:52

---

Data Download Date/Time: 05/30/2018 11:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PECK,KEITH,D  
BREATH TEST OPERATOR AGENCY: Volusia County SO  
ARRESTING OFFICER: LIPPACK,KARL  
ARRESTING OFFICER AGENCY: Ormond Beach PD  
LAST AGENCY INSPECTION: 04/16/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 644476  
DRY GAS EXPIRY DATE: 12/01/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: SNELL

DATE/TIME OF TEST: 04/30/2018 03:18:17

---

Diagnostics Check	OK	03:10
Air Blank	0.000	03:11
Control Test	0.081	03:11
Air Blank	0.000	03:12
Subject Sample # 1	0.000	03:13
Breath Volume	1.308	03:13
Air Blank	0.000	03:13
Air Blank	0.000	03:15
Subject Sample # 2	0.000	03:16
Breath Volume	1.210	03:16
Air Blank	0.000	03:17
Control Test	0.081	03:17
Air Blank	0.000	03:18
Diagnostics Check	OK	03:18

---

Data Download Date/Time: 05/30/2018 11:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001155

INSTRUMENT REGISTERED TO: EGLIN AFB  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MATHRE,PETER,M  
BREATH TEST OPERATOR AGENCY: Eglin AFB PD  
ARRESTING OFFICER: BYLICKI,DANIEL  
ARRESTING OFFICER AGENCY: Eglin AFB PD  
LAST AGENCY INSPECTION: 03/08/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A5  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 00:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: PAGE

DATE/TIME OF TEST: 04/01/2018 01:33:53

---

Diagnostics Check	OK	01:27
Air Blank	0.000	01:27
Control Test	0.080	01:28
Air Blank	0.000	01:28
Subject Sample # 1	0.000	01:29
Breath Volume	2.253	01:29
Air Blank	0.000	01:30
Air Blank	0.000	01:32
Subject Sample # 2	0.000	01:32
Breath Volume	2.968	01:32
Air Blank	0.000	01:33
Control Test	0.080	01:33
Air Blank	0.000	01:33
Diagnostics Check	OK	01:33

---

Data Download Date/Time: 04/30/2018 14:35



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001155

INSTRUMENT REGISTERED TO: EGLIN AFB  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MATHRE,PETER,M  
BREATH TEST OPERATOR AGENCY: Eglin AFB PD  
ARRESTING OFFICER: CAMPBELL,STEVEN  
ARRESTING OFFICER AGENCY: Eglin AFB PD  
LAST AGENCY INSPECTION: 03/08/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A5  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 22:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: RUMBLE

DATE/TIME OF TEST: 04/05/2018 23:14:30

---

Diagnostics Check	OK	23:08
Air Blank	0.000	23:08
Control Test	0.080	23:08
Air Blank	0.000	23:09
Subject Sample # 1	0.233	23:10
Breath Volume	1.328	23:10
Air Blank	0.000	23:10
Air Blank	0.000	23:12
Subject Sample # 2	0.240	23:12
Breath Volume	2.242	23:12
Air Blank	0.000	23:13
Control Test	0.079	23:13
Air Blank	0.000	23:14
Diagnostics Check	OK	23:14

---

Data Download Date/Time: 04/30/2018 14:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RUBIN,RICHARD,D  
BREATH TEST OPERATOR AGENCY: Ormond Beach PD  
ARRESTING OFFICER: RAIHALA,J  
ARRESTING OFFICER AGENCY: Ormond Beach PD  
LAST AGENCY INSPECTION: 03/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814183  
DRY GAS EXPIRY DATE: 07/26/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: SARZIER

DATE/TIME OF TEST: 04/01/2018 02:22:44

---

Diagnostics Check	OK	02:19	
Air Blank	0.000	02:20	
Control Test	0.080	02:20	
Air Blank	0.000	02:21	
Subject Sample	# 1	02:21	Subject Test Refused
Breath Volume	0.000	02:21	
Air Blank	0.000	02:21	
Control Test	0.080	02:22	
Air Blank	0.000	02:22	
Diagnostics Check	OK	02:22	

---

Data Download Date/Time: 04/29/2018 08:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RUBIN,RICHARD,D  
BREATH TEST OPERATOR AGENCY: Ormond Beach PD  
ARRESTING OFFICER: RAIHALA,J  
ARRESTING OFFICER AGENCY: Ormond Beach PD  
LAST AGENCY INSPECTION: 03/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814183  
DRY GAS EXPIRY DATE: 07/26/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: SARZIER

DATE/TIME OF TEST: 04/01/2018 03:09:07

---

Diagnostics Check	OK	03:01
Air Blank	0.000	03:02
Control Test	0.080	03:02
Air Blank	0.000	03:03
Subject Sample # 1	0.180	03:04
Breath Volume	1.109	03:04
Air Blank	0.000	03:04
Air Blank	0.000	03:06
Subject Sample # 2	0.174	03:07
Breath Volume	1.128	03:07
Air Blank	0.000	03:08
Control Test	0.080	03:08
Air Blank	0.000	03:09
Diagnostics Check	OK	03:09

---

Data Download Date/Time: 04/29/2018 08:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RUBIN,RICHARD,D  
BREATH TEST OPERATOR AGENCY: Ormond Beach PD  
ARRESTING OFFICER: RAIHALA,J  
ARRESTING OFFICER AGENCY: Ormond Beach PD  
LAST AGENCY INSPECTION: 03/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814183  
DRY GAS EXPIRY DATE: 07/26/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:28  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: HARRINGTON

DATE/TIME OF TEST: 04/02/2018 04:14:19

---

Diagnostics Check	OK	04:02	
Air Blank	0.000	04:03	
Control Test	0.081	04:03	
Air Blank	0.000	04:04	
Subject Sample	# 1	04:07	No Sample Provided
Breath Volume	0.000	04:07	
Air Blank	0.000	04:07	
Air Blank	0.000	04:09	
Subject Sample	# 2	04:12	No Sample Provided
Breath Volume	0.000	04:12	
Air Blank	0.000	04:13	
Control Test	0.081	04:13	
Air Blank	0.000	04:14	
Diagnostics Check	OK	04:14	

---

Data Download Date/Time: 04/29/2018 08:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RUBIN,RICHARD,D  
BREATH TEST OPERATOR AGENCY: Ormond Beach PD  
ARRESTING OFFICER: RAIHALA,J  
ARRESTING OFFICER AGENCY: Ormond Beach PD  
LAST AGENCY INSPECTION: 03/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814183  
DRY GAS EXPIRY DATE: 07/26/2020  
VIOLATION CODE: DUI  
ARREST TIME: 19:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:36

SUBJECT'S LAST NAME: CRICK

DATE/TIME OF TEST: 04/09/2018 21:02:36

---

Diagnostics Check	OK	20:59	
Air Blank	0.000	20:59	
Control Test	0.081	21:00	
Air Blank	0.000	21:00	
Subject Sample	# 1	21:01	Subject Test Refused
Breath Volume	0.000	21:01	
Air Blank	0.000	21:01	
Control Test	0.081	21:02	
Air Blank	0.000	21:02	
Diagnostics Check	OK	21:02	

---

Data Download Date/Time: 04/29/2018 08:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RUBIN,RICHARD,D  
BREATH TEST OPERATOR AGENCY: Ormond Beach PD  
ARRESTING OFFICER: RAIHALA,J  
ARRESTING OFFICER AGENCY: Ormond Beach PD  
LAST AGENCY INSPECTION: 03/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814183  
DRY GAS EXPIRY DATE: 07/26/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: HIGGINS

DATE/TIME OF TEST: 04/11/2018 00:02:18

---

Diagnostics Check	OK	23:56
Air Blank	0.000	23:56
Control Test	0.080	23:56
Air Blank	0.000	23:57
Subject Sample # 2	0.227	23:57
Breath Volume	1.761	23:57
Air Blank	0.000	23:58
Air Blank	0.000	00:00
Subject Sample # 1	0.218	00:00
Breath Volume	2.519	00:00
Air Blank	0.000	00:01
Control Test	0.080	00:01
Air Blank	0.000	00:02
Diagnostics Check	OK	00:02

---

Data Download Date/Time: 04/29/2018 08:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARNER, MICHAEL, D  
BREATH TEST OPERATOR AGENCY: Ormond Beach PD  
ARRESTING OFFICER: LLANES, RAMON  
ARRESTING OFFICER AGENCY: Ormond Beach PD  
LAST AGENCY INSPECTION: 03/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814183  
DRY GAS EXPIRY DATE: 07/26/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 04/12/2018 03:07:33

---

Diagnostics Check		OK	03:00
Air Blank		0.000	03:01
Control Test		0.081	03:01
Air Blank		0.000	03:01
Subject Sample	# 1	0.188	03:03
Breath Volume		1.296	03:03
Air Blank		0.000	03:03
Air Blank		0.000	03:05
Subject Sample	# 2	0.183	03:06
Breath Volume		1.960	03:06
Air Blank		0.000	03:06
Control Test		0.081	03:06
Air Blank		0.000	03:07
Diagnostics Check		OK	03:07

---

Data Download Date/Time: 04/29/2018 08:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RUBIN,RICHARD,D  
BREATH TEST OPERATOR AGENCY: Ormond Beach PD  
ARRESTING OFFICER: RAIHALA,J  
ARRESTING OFFICER AGENCY: Ormond Beach PD  
LAST AGENCY INSPECTION: 03/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814183  
DRY GAS EXPIRY DATE: 07/26/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:38  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: STARKER

DATE/TIME OF TEST: 04/16/2018 03:56:04

---

Diagnostics Check	OK	03:48
Air Blank	0.000	03:48
Control Test	0.080	03:49
Air Blank	0.000	03:49
Subject Sample # 1	0.188	03:50
Breath Volume	1.839	03:50
Air Blank	0.000	03:51
Air Blank	0.000	03:52
Subject Sample # 2	0.185	03:54
Breath Volume	1.500	03:54
Air Blank	0.000	03:55
Control Test	0.080	03:55
Air Blank	0.000	03:55
Diagnostics Check	OK	03:56

---

Data Download Date/Time: 04/29/2018 08:10



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAMILTON,BENITA,L  
BREATH TEST OPERATOR AGENCY: Ormond Beach PD  
ARRESTING OFFICER: KAVANAUGH,ANTHONY  
ARRESTING OFFICER AGENCY: Ormond Beach PD  
LAST AGENCY INSPECTION: 03/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814183  
DRY GAS EXPIRY DATE: 07/26/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: BURKE

DATE/TIME OF TEST: 04/22/2018 04:25:36

---

Diagnostics Check	OK	04:19
Air Blank	0.000	04:19
Control Test	0.080	04:19
Air Blank	0.000	04:20
Subject Sample # 1	0.000	04:21
Breath Volume	2.195	04:21
Air Blank	0.000	04:21
Air Blank	0.000	04:23
Subject Sample # 2	0.000	04:24
Breath Volume	1.664	04:24
Air Blank	0.000	04:24
Control Test	0.079	04:25
Air Blank	0.000	04:25
Diagnostics Check	OK	04:25

---

Data Download Date/Time: 04/29/2018 08:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WARMINGTON, WILLIAM, G  
BREATH TEST OPERATOR AGENCY: Ormond Beach PD  
ARRESTING OFFICER: GALLAGHER, JOSEPH  
ARRESTING OFFICER AGENCY: Ormond Beach PD  
LAST AGENCY INSPECTION: 03/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814183  
DRY GAS EXPIRY DATE: 07/26/2020  
VIOLATION CODE: DUI  
ARREST TIME: 14:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 15:35

SUBJECT'S LAST NAME: STANLEY

DATE/TIME OF TEST: 04/23/2018 16:01:25

---

Diagnostics Check	OK	15:58	
Air Blank	0.000	15:59	
Control Test	0.081	15:59	
Air Blank	0.000	15:59	
Subject Sample	# 1	16:00	Subject Test Refused
Breath Volume	0.000	16:00	
Air Blank	0.000	16:00	
Control Test	0.080	16:00	
Air Blank	0.000	16:01	
Diagnostics Check	OK	16:01	

---

Data Download Date/Time: 04/29/2018 08:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001159

INSTRUMENT REGISTERED TO: ORANGE CITY P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYERS, MATTHEW, A  
BREATH TEST OPERATOR AGENCY: Volusia County SO  
ARRESTING OFFICER: MYERS, MATTHEW  
ARRESTING OFFICER AGENCY: Volusia County SO  
LAST AGENCY INSPECTION: 03/16/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17316080A2  
DRY GAS EXPIRY DATE: 08/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 16:03  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 16:20

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 04/07/2018 16:47:12

---

Diagnostics Check	OK	16:40
Air Blank	0.000	16:41
Control Test	0.080	16:41
Air Blank	0.000	16:41
Subject Sample # 1	0.165	16:42
Breath Volume	1.839	16:42
Air Blank	0.000	16:43
Air Blank	0.000	16:44
Subject Sample # 2	0.168	16:45
Breath Volume	1.867	16:45
Air Blank	0.000	16:46
Control Test	0.080	16:46
Air Blank	0.000	16:47
Diagnostics Check	OK	16:47

---

Data Download Date/Time: 04/16/2018 19:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001159

INSTRUMENT REGISTERED TO: ORANGE CITY P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PILARSKI,BRIAN,J  
BREATH TEST OPERATOR AGENCY: Orange City PD  
ARRESTING OFFICER: SMITH,BRIE  
ARRESTING OFFICER AGENCY: Orange City PD  
LAST AGENCY INSPECTION: 03/16/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17316080A2  
DRY GAS EXPIRY DATE: 08/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 04:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: RIEL 3

DATE/TIME OF TEST: 04/10/2018 06:00:41

---

Diagnostics Check	OK	05:54
Air Blank	0.000	05:54
Control Test	0.081	05:55
Air Blank	0.000	05:55
Subject Sample # 1	0.000	05:56
Breath Volume	3.378	05:56
Air Blank	0.000	05:56
Air Blank	0.000	05:58
Subject Sample # 2	0.000	05:59
Breath Volume	3.390	05:59
Air Blank	0.000	05:59
Control Test	0.081	06:00
Air Blank	0.000	06:00
Diagnostics Check	OK	06:00

---

Data Download Date/Time: 04/16/2018 19:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001159

INSTRUMENT REGISTERED TO: ORANGE CITY P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRUBAKER,KELLY,J  
BREATH TEST OPERATOR AGENCY: Orange City PD  
ARRESTING OFFICER: CHARLES,COREY  
ARRESTING OFFICER AGENCY: Volusia County SO  
LAST AGENCY INSPECTION: 04/16/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17316080A2  
DRY GAS EXPIRY DATE: 08/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 00:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: PATRICK III

DATE/TIME OF TEST: 04/20/2018 00:59:54

---

Diagnostics Check	OK	00:58	
Air Blank		00:59	Ambient Fail
Air Blank	0.000	00:59	

---

Data Download Date/Time: 05/15/2018 16:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001159

INSTRUMENT REGISTERED TO: ORANGE CITY P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRUBAKER,KELLY,J  
BREATH TEST OPERATOR AGENCY: Orange City PD  
ARRESTING OFFICER: CHARLES,COREY  
ARRESTING OFFICER AGENCY: Volusia County SO  
LAST AGENCY INSPECTION: 04/16/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17316080A2  
DRY GAS EXPIRY DATE: 08/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 00:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: PATRICK III

DATE/TIME OF TEST: 04/20/2018 01:09:19

---

Diagnostics Check		OK	01:02
Air Blank		0.000	01:03
Control Test		0.079	01:03
Air Blank		0.000	01:04
Subject Sample	# 1	0.164	01:04
Breath Volume		1.863	01:04
Air Blank		0.000	01:05
Air Blank		0.000	01:07
Subject Sample	# 2	0.171	01:07
Breath Volume		3.093	01:07
Air Blank		0.000	01:08
Control Test		0.078	01:08
Air Blank		0.000	01:09
Diagnostics Check		OK	01:09

---

Data Download Date/Time: 05/15/2018 16:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

INSTRUMENT REGISTERED TO: LADY LAKE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DURYEAMATTHEW,W  
BREATH TEST OPERATOR AGENCY: Lady Lake PD  
ARRESTING OFFICER: DURYEAMATTHEW  
ARRESTING OFFICER AGENCY: Lady Lake PD  
LAST AGENCY INSPECTION: 03/14/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17817080A3  
DRY GAS EXPIRY DATE: 08/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 19:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:42

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 04/07/2018 20:15:01

---

Diagnostics Check	OK	20:11	
Air Blank	0.000	20:12	
Control Test	0.076	20:12	
Air Blank	0.000	20:13	
Subject Sample	# 1	20:13	Subject Test Refused
Breath Volume	0.000	20:13	
Air Blank	0.000	20:14	
Control Test	0.076	20:14	
Air Blank	0.000	20:14	
Diagnostics Check	OK	20:15	

---

Data Download Date/Time: 04/16/2018 08:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001190

INSTRUMENT REGISTERED TO: BOYNTON BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CASTRO,DENNIS,  
BREATH TEST OPERATOR AGENCY: Boynton Beach PD  
ARRESTING OFFICER: CASTRO,DENNIS  
ARRESTING OFFICER AGENCY: Boynton Beach PD  
LAST AGENCY INSPECTION: 04/01/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 390314  
DRY GAS EXPIRY DATE: 04/19/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: GRANADOS OLIVA

DATE/TIME OF TEST: 04/04/2018 22:26:03

---

Diagnostics Check	OK	22:21	
Air Blank	0.000	22:22	
Control Test	0.080	22:22	
Air Blank	0.000	22:22	
Subject Sample	# 1	22:24	Subject Test Refused
Breath Volume	0.000	22:24	
Air Blank	0.000	22:25	
Control Test	0.080	22:25	
Air Blank	0.000	22:25	
Diagnostics Check	OK	22:26	

---

Data Download Date/Time: 04/28/2018 18:39



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001190

INSTRUMENT REGISTERED TO: BOYNTON BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CASTRO,DENNIS,  
BREATH TEST OPERATOR AGENCY: Boynton Beach PD  
ARRESTING OFFICER: CASTRO,DENNIS  
ARRESTING OFFICER AGENCY: Boynton Beach PD  
LAST AGENCY INSPECTION: 04/01/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 390314  
DRY GAS EXPIRY DATE: 04/19/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: PIGEOT

DATE/TIME OF TEST: 04/07/2018 03:10:43

---

Diagnostics Check	OK	03:04
Air Blank	0.000	03:04
Control Test	0.080	03:04
Air Blank	0.000	03:05
Subject Sample # 1	0.210	03:06
Breath Volume	1.613	03:06
Air Blank	0.000	03:06
Air Blank	0.000	03:08
Subject Sample # 2	0.220	03:09
Breath Volume	2.562	03:09
Air Blank	0.000	03:09
Control Test	0.079	03:10
Air Blank	0.000	03:10
Diagnostics Check	OK	03:10

---

Data Download Date/Time: 04/28/2018 18:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001190

INSTRUMENT REGISTERED TO: BOYNTON BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CASTRO,DENNIS,  
BREATH TEST OPERATOR AGENCY: Boynton Beach PD  
ARRESTING OFFICER: CASTRO,DENNIS  
ARRESTING OFFICER AGENCY: Boynton Beach PD  
LAST AGENCY INSPECTION: 04/01/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 390314  
DRY GAS EXPIRY DATE: 04/19/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: MOONEY

DATE/TIME OF TEST: 04/27/2018 00:01:20

---

Diagnostics Check	OK	23:55
Air Blank	0.000	23:55
Control Test	0.081	23:56
Air Blank	0.000	23:56
Subject Sample # 1	0.290	23:56
Breath Volume	2.179	23:56
Air Blank	0.000	23:57
Air Blank	0.000	23:59
Subject Sample # 2	0.293	23:59
Breath Volume	2.019	23:59
Air Blank	0.000	00:00
Control Test	0.080	00:00
Air Blank	0.000	00:01
Diagnostics Check	OK	00:01

---

Data Download Date/Time: 04/28/2018 18:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LOPEZ,JOSE,D  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: LOPEZ,JOSE  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17316080A3  
DRY GAS EXPIRY DATE: 08/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 21:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:53

SUBJECT'S LAST NAME: PARRENO GONZALEZ

DATE/TIME OF TEST: 04/17/2018 22:24:14

---

Diagnostics Check	OK	22:17
Air Blank	0.000	22:17
Control Test	0.079	22:18
Air Blank	0.000	22:18
Subject Sample # 1	0.150	22:19
Breath Volume	1.132	22:19
Air Blank	0.000	22:19
Air Blank	0.000	22:21
Subject Sample # 2	0.164	22:22
Breath Volume	3.527	22:22
Air Blank	0.000	22:23
Control Test	0.078	22:23
Air Blank	0.000	22:24
Diagnostics Check	OK	22:24

---

Data Download Date/Time: 05/01/2018 17:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LOPEZ,JOSE,D  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: LOPEZ,JOSE  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17316080A3  
DRY GAS EXPIRY DATE: 08/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 18:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:11

SUBJECT'S LAST NAME: GARCIA FIGUEROA

DATE/TIME OF TEST: 04/20/2018 19:47:53

---

Diagnostics Check	OK	19:36	
Air Blank	0.000	19:36	
Control Test	0.079	19:37	
Air Blank	0.000	19:37	
Subject Sample	# 1 0.165	19:40	Volume Not Met
Breath Volume	0.066	19:40	
Air Blank	0.000	19:41	
Air Blank	0.000	19:43	
Subject Sample	# 2	19:46	Subject Test Refused
Breath Volume	0.000	19:46	
Air Blank	0.000	19:46	
Control Test	0.077	19:47	
Air Blank	0.000	19:47	
Diagnostics Check	OK	19:47	

---

Data Download Date/Time: 05/01/2018 17:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LOPEZ,JOSE,D  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: LOPEZ,JOSE  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17316080A3  
DRY GAS EXPIRY DATE: 08/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 23:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: CANTU FLORES

DATE/TIME OF TEST: 04/21/2018 23:38:26

---

Diagnostics Check	OK	23:31
Air Blank	0.000	23:32
Control Test	0.077	23:32
Air Blank	0.000	23:32
Subject Sample # 1	0.167	23:33
Breath Volume	2.257	23:33
Air Blank	0.000	23:34
Air Blank	0.000	23:36
Subject Sample # 2	0.164	23:36
Breath Volume	2.433	23:36
Air Blank	0.000	23:37
Control Test	0.077	23:37
Air Blank	0.000	23:38
Diagnostics Check	OK	23:38

---

Data Download Date/Time: 05/01/2018 17:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THORE,RICHARD,A  
BREATH TEST OPERATOR AGENCY: Panama City PD  
ARRESTING OFFICER: THORE,RICHARD  
ARRESTING OFFICER AGENCY: Panama City PD  
LAST AGENCY INSPECTION: 03/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 366849  
DRY GAS EXPIRY DATE: 03/30/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 04/03/2018 00:30:19

---

Diagnostics Check	OK	00:27	
Air Blank	0.000	00:28	
Control Test	0.080	00:28	
Air Blank	0.000	00:28	
Subject Sample	# 1	00:28	Subject Test Refused
Breath Volume	0.000	00:28	
Air Blank	0.000	00:29	
Control Test	0.080	00:29	
Air Blank	0.000	00:30	
Diagnostics Check	OK	00:30	

---

Data Download Date/Time: 04/12/2018 02:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ABRAHAM,ERIC,D  
BREATH TEST OPERATOR AGENCY: Panama City PD  
ARRESTING OFFICER: ABRAHAM,ERIC  
ARRESTING OFFICER AGENCY: Panama City PD  
LAST AGENCY INSPECTION: 03/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 366849  
DRY GAS EXPIRY DATE: 03/30/2019  
VIOLATION CODE: DUI  
ARREST TIME: 16:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:41

SUBJECT'S LAST NAME: GALLAGHER

DATE/TIME OF TEST: 04/03/2018 17:11:06

---

Diagnostics Check		OK	17:01	
Air Blank		0.000	17:02	
Control Test		0.079	17:02	
Air Blank		0.000	17:02	
Subject Sample	# 1	0.235	17:06	Volume Not Met
Breath Volume		0.921	17:06	
Air Blank		0.000	17:06	
Air Blank		0.000	17:08	
Subject Sample	# 2		17:09	Subject Test Refused
Breath Volume		0.000	17:09	
Air Blank		0.000	17:10	
Control Test		0.078	17:10	
Air Blank		0.000	17:11	
Diagnostics Check		OK	17:11	

---

Data Download Date/Time: 04/12/2018 02:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FLORES,BARRY,G  
BREATH TEST OPERATOR AGENCY: Panama City PD  
ARRESTING OFFICER: FLORES,BARRY  
ARRESTING OFFICER AGENCY: Panama City PD  
LAST AGENCY INSPECTION: 03/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 366849  
DRY GAS EXPIRY DATE: 03/30/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: OWENS

DATE/TIME OF TEST: 04/06/2018 03:19:14

---

Diagnostics Check	OK	03:12
Air Blank	0.000	03:13
Control Test	0.080	03:13
Air Blank	0.000	03:14
Subject Sample # 1	0.089	03:14
Breath Volume	2.371	03:14
Air Blank	0.000	03:15
Air Blank	0.000	03:17
Subject Sample # 2	0.082	03:17
Breath Volume	2.464	03:17
Air Blank	0.000	03:18
Control Test	0.079	03:18
Air Blank	0.000	03:19
Diagnostics Check	OK	03:19

---

Data Download Date/Time: 04/12/2018 02:41



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAMSEY,JOHNATHAN,L  
BREATH TEST OPERATOR AGENCY: Panama City PD  
ARRESTING OFFICER: RHODES,WILLIAM  
ARRESTING OFFICER AGENCY: Panama City PD  
LAST AGENCY INSPECTION: 04/11/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 366849  
DRY GAS EXPIRY DATE: 03/30/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 04/15/2018 00:16:35

---

Diagnostics Check		OK	00:04	
Air Blank		0.000	00:05	
Control Test		0.080	00:05	
Air Blank		0.000	00:05	
Subject Sample	# 1	0.000	00:09	Volume Not Met
Breath Volume		0.277	00:09	
Air Blank		0.000	00:09	
Air Blank		0.000	00:11	
Subject Sample	# 2	0.000	00:12	
Breath Volume		2.417	00:12	
Air Blank		0.000	00:12	
Air Blank		0.000	00:14	
Subject Sample	# 3	0.000	00:15	
Breath Volume		1.371	00:15	
Air Blank		0.000	00:15	
Control Test		0.080	00:15	
Air Blank		0.000	00:16	
Diagnostics Check		OK	00:16	

---

Data Download Date/Time: 05/28/2018 10:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALLYN,WILLIAM,P  
BREATH TEST OPERATOR AGENCY: Panama City PD  
ARRESTING OFFICER: ALLYN,WILLIAM  
ARRESTING OFFICER AGENCY: Panama City PD  
LAST AGENCY INSPECTION: 04/11/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 366849  
DRY GAS EXPIRY DATE: 03/30/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: ISAACS

DATE/TIME OF TEST: 04/15/2018 22:56:42

---

Diagnostics Check	OK	22:50
Air Blank	0.000	22:51
Control Test	0.080	22:51
Air Blank	0.000	22:51
Subject Sample # 1	0.178	22:52
Breath Volume	3.960	22:52
Air Blank	0.000	22:52
Air Blank	0.000	22:54
Subject Sample # 2	0.175	22:55
Breath Volume	3.984	22:55
Air Blank	0.000	22:55
Control Test	0.080	22:56
Air Blank	0.000	22:56
Diagnostics Check	OK	22:56

---

Data Download Date/Time: 05/28/2018 10:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MERRITT, DAKOTA, A  
BREATH TEST OPERATOR AGENCY: Panama City PD  
ARRESTING OFFICER: LEATH, JOHNATHAN  
ARRESTING OFFICER AGENCY: Panama City PD  
LAST AGENCY INSPECTION: 04/11/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 366849  
DRY GAS EXPIRY DATE: 03/30/2019  
VIOLATION CODE: DUI  
ARREST TIME: 19:59  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:16

SUBJECT'S LAST NAME: CHERRY

DATE/TIME OF TEST: 04/26/2018 20:42:54

---

Diagnostics Check	OK	20:36
Air Blank	0.000	20:37
Control Test	0.080	20:37
Air Blank	0.000	20:37
Subject Sample # 1	0.167	20:38
Breath Volume	1.597	20:38
Air Blank	0.000	20:38
Air Blank	0.000	20:40
Subject Sample # 2	0.175	20:41
Breath Volume	1.667	20:41
Air Blank	0.000	20:42
Control Test	0.079	20:42
Air Blank	0.000	20:42
Diagnostics Check	OK	20:42

---

Data Download Date/Time: 05/28/2018 10:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FLORES,BARRY,G  
BREATH TEST OPERATOR AGENCY: Panama City PD  
ARRESTING OFFICER: FLORES,BARRY  
ARRESTING OFFICER AGENCY: Panama City PD  
LAST AGENCY INSPECTION: 04/11/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 366849  
DRY GAS EXPIRY DATE: 03/30/2019  
VIOLATION CODE: DUI  
ARREST TIME: 06:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:25

SUBJECT'S LAST NAME: MAXWELL

DATE/TIME OF TEST: 04/28/2018 06:51:27

---

Diagnostics Check	OK	06:45
Air Blank	0.000	06:45
Control Test	0.080	06:46
Air Blank	0.000	06:46
Subject Sample # 1	0.000	06:47
Breath Volume	2.324	06:47
Air Blank	0.000	06:47
Air Blank	0.000	06:49
Subject Sample # 2	0.000	06:50
Breath Volume	1.988	06:50
Air Blank	0.000	06:50
Control Test	0.079	06:50
Air Blank	0.000	06:51
Diagnostics Check	OK	06:51

---

Data Download Date/Time: 05/28/2018 10:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TURNER,JONATHAN,R  
BREATH TEST OPERATOR AGENCY: Panama City PD  
ARRESTING OFFICER: TURNER,JONATHAN  
ARRESTING OFFICER AGENCY: Panama City PD  
LAST AGENCY INSPECTION: 04/11/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 366849  
DRY GAS EXPIRY DATE: 03/30/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:12  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:46

SUBJECT'S LAST NAME: COWAN

DATE/TIME OF TEST: 04/28/2018 23:10:52

---

Diagnostics Check	OK	23:07	
Air Blank	0.000	23:07	
Control Test	0.080	23:08	
Air Blank	0.000	23:08	
Subject Sample	# 1	23:09	Subject Test Refused
Breath Volume	0.000	23:09	
Air Blank	0.000	23:09	
Control Test	0.080	23:10	
Air Blank	0.000	23:10	
Diagnostics Check	OK	23:10	

---

Data Download Date/Time: 05/28/2018 10:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

INSTRUMENT REGISTERED TO: TYNDALL AFB  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSON,CHRISTOPHER,R  
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD  
ARRESTING OFFICER: JOHNSON,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Tyndall AFB PD  
LAST AGENCY INSPECTION: 04/04/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32316080A3  
DRY GAS EXPIRY DATE: 01/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 04/14/2018 03:26:49

---

Diagnostics Check	OK	03:20
Air Blank	0.000	03:21
Control Test	0.080	03:21
Air Blank	0.000	03:22
Subject Sample # 1	0.195	03:22
Breath Volume	4.183	03:22
Air Blank	0.000	03:23
Air Blank	0.000	03:24
Subject Sample # 2	0.191	03:25
Breath Volume	3.707	03:25
Air Blank	0.000	03:25
Control Test	0.080	03:26
Air Blank	0.000	03:26
Diagnostics Check	OK	03:26

---

Data Download Date/Time: 05/10/2018 15:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

INSTRUMENT REGISTERED TO: TYNDALL AFB  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSON,CHRISTOPHER,R  
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD  
ARRESTING OFFICER: JOHNSON,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Tyndall AFB PD  
LAST AGENCY INSPECTION: 04/04/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32316080A3  
DRY GAS EXPIRY DATE: 01/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 05:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:24

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 04/15/2018 05:50:39

---

Diagnostics Check	OK	05:44
Air Blank	0.000	05:44
Control Test	0.080	05:45
Air Blank	0.000	05:45
Subject Sample # 1	0.253	05:46
Breath Volume	3.382	05:46
Air Blank	0.000	05:46
Air Blank	0.000	05:48
Subject Sample # 2	0.266	05:49
Breath Volume	4.210	05:49
Air Blank	0.000	05:49
Control Test	0.079	05:50
Air Blank	0.000	05:50
Diagnostics Check	OK	05:50

---

Data Download Date/Time: 05/10/2018 15:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

INSTRUMENT REGISTERED TO: TYNDALL AFB  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PLACE,NICHOLAS,T  
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD  
ARRESTING OFFICER: FIGUEROA,JENILY  
ARRESTING OFFICER AGENCY: Tyndall AFB PD  
LAST AGENCY INSPECTION: 04/04/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32316080A3  
DRY GAS EXPIRY DATE: 01/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: TODD

DATE/TIME OF TEST: 04/22/2018 02:05:05

---

Diagnostics Check	OK	01:58
Air Blank	0.000	01:59
Control Test	0.080	01:59
Air Blank	0.000	02:00
Subject Sample # 1	0.060	02:00
Breath Volume	1.902	02:00
Air Blank	0.000	02:01
Air Blank	0.000	02:03
Subject Sample # 2	0.056	02:03
Breath Volume	1.214	02:03
Air Blank	0.000	02:04
Control Test	0.080	02:04
Air Blank	0.000	02:05
Diagnostics Check	OK	02:05

---

Data Download Date/Time: 05/10/2018 15:21



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

INSTRUMENT REGISTERED TO: TYNDALL AFB  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PLACE,NICHOLAS,T  
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD  
ARRESTING OFFICER: HOWARD,WENDELL  
ARRESTING OFFICER AGENCY: Tyndall AFB PD  
LAST AGENCY INSPECTION: 04/04/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32316080A3  
DRY GAS EXPIRY DATE: 01/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: HUNT

DATE/TIME OF TEST: 04/22/2018 03:34:08

---

Diagnostics Check		OK	03:21	
Air Blank		0.000	03:22	
Control Test		0.080	03:22	
Air Blank		0.000	03:23	
Subject Sample	# 1	0.045	03:26	Volume Not Met
Breath Volume		0.937	03:26	
Air Blank		0.000	03:26	
Air Blank		0.000	03:28	
Subject Sample	# 2	0.049	03:29	
Breath Volume		1.734	03:29	
Air Blank		0.000	03:30	
Air Blank		0.000	03:31	
Subject Sample	# 3	0.048	03:32	
Breath Volume		1.300	03:32	
Air Blank		0.000	03:33	
Control Test		0.081	03:33	
Air Blank		0.000	03:34	
Diagnostics Check		OK	03:34	

---

Data Download Date/Time: 05/10/2018 15:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

INSTRUMENT REGISTERED TO: TYNDALL AFB  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSON,CHRISTOPHER,R  
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD  
ARRESTING OFFICER: JOHNSON,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Tyndall AFB PD  
LAST AGENCY INSPECTION: 04/04/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32316080A3  
DRY GAS EXPIRY DATE: 01/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: ALEXANDER

DATE/TIME OF TEST: 04/28/2018 02:02:18

---

Diagnostics Check	OK	01:55
Air Blank	0.000	01:55
Control Test	0.080	01:56
Air Blank	0.000	01:56
Subject Sample # 1	0.155	01:57
Breath Volume	1.824	01:57
Air Blank	0.000	01:57
Air Blank	0.000	01:59
Subject Sample # 2	0.144	02:00
Breath Volume	1.429	02:00
Air Blank	0.000	02:01
Control Test	0.079	02:01
Air Blank	0.000	02:02
Diagnostics Check	OK	02:02

---

Data Download Date/Time: 05/10/2018 15:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

INSTRUMENT REGISTERED TO: TYNDALL AFB  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSON,CHRISTOPHER,R  
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD  
ARRESTING OFFICER: JOHNSON,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Tyndall AFB PD  
LAST AGENCY INSPECTION: 04/04/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32316080A3  
DRY GAS EXPIRY DATE: 01/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: FLOYD

DATE/TIME OF TEST: 04/29/2018 03:10:56

---

Diagnostics Check	OK	03:07	
Air Blank	0.000	03:08	
Control Test	0.078	03:08	
Air Blank	0.000	03:09	
Subject Sample	# 1	03:09	Subject Test Refused
Breath Volume	0.000	03:09	
Air Blank	0.000	03:10	
Control Test	0.079	03:10	
Air Blank	0.000	03:10	
Diagnostics Check	OK	03:10	

---

Data Download Date/Time: 05/10/2018 15:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMAS,MICAH,L  
BREATH TEST OPERATOR AGENCY: West Palm Beach PD  
ARRESTING OFFICER: THOMAS,MICAH  
ARRESTING OFFICER AGENCY: West Palm Beach PD  
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 152169  
DRY GAS EXPIRY DATE: 10/30/2018  
VIOLATION CODE: DUI  
ARREST TIME: 22:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: KIMBROUGH

DATE/TIME OF TEST: 04/05/2018 23:30:05

---

Diagnostics Check	OK	23:21
Air Blank	0.000	23:21
Control Test	0.080	23:21
Air Blank	0.000	23:22
Subject Sample # 1	0.199	23:23
Breath Volume	1.265	23:23
Air Blank	0.000	23:24
Air Blank	0.000	23:26
Subject Sample # 2	0.197	23:28
Breath Volume	1.175	23:28
Air Blank	0.000	23:29
Control Test	0.079	23:29
Air Blank	0.000	23:29
Diagnostics Check	OK	23:30

---

Data Download Date/Time: 04/29/2018 08:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ANNESE,MORRIS,J  
BREATH TEST OPERATOR AGENCY: West Palm Beach PD  
ARRESTING OFFICER: ANNESE,MORRIS  
ARRESTING OFFICER AGENCY: West Palm Beach PD  
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 152169  
DRY GAS EXPIRY DATE: 10/30/2018  
VIOLATION CODE: DUI  
ARREST TIME: 21:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: MIGUEL ESCALANTE

DATE/TIME OF TEST: 04/14/2018 22:18:12

---

Diagnostics Check	OK	22:15	
Air Blank	0.000	22:15	
Control Test	0.080	22:16	
Air Blank	0.000	22:16	
Subject Sample	# 1	22:16	Subject Test Refused
Breath Volume	0.000	22:16	
Air Blank	0.000	22:17	
Control Test	0.079	22:17	
Air Blank	0.000	22:18	
Diagnostics Check	OK	22:18	

---

Data Download Date/Time: 04/29/2018 08:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ANNESE,MORRIS,J  
BREATH TEST OPERATOR AGENCY: West Palm Beach PD  
ARRESTING OFFICER: ANNESE,MORRIS  
ARRESTING OFFICER AGENCY: West Palm Beach PD  
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 152169  
DRY GAS EXPIRY DATE: 10/30/2018  
VIOLATION CODE: DUI  
ARREST TIME: 04:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: MIGUEL ROBLERO

DATE/TIME OF TEST: 04/15/2018 05:45:01

---

Diagnostics Check	OK	05:39
Air Blank	0.000	05:39
Control Test	0.080	05:39
Air Blank	0.000	05:40
Subject Sample # 1	0.162	05:40
Breath Volume	2.996	05:40
Air Blank	0.000	05:41
Air Blank	0.000	05:43
Subject Sample # 2	0.163	05:43
Breath Volume	2.613	05:43
Air Blank	0.000	05:44
Control Test	0.080	05:44
Air Blank	0.000	05:44
Diagnostics Check	OK	05:45

---

Data Download Date/Time: 04/29/2018 08:00

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ANNESE,MORRIS,J  
BREATH TEST OPERATOR AGENCY: West Palm Beach PD  
ARRESTING OFFICER: ANNESE,MORRIS  
ARRESTING OFFICER AGENCY: West Palm Beach PD  
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 152169  
DRY GAS EXPIRY DATE: 10/30/2018  
VIOLATION CODE: DUI  
ARREST TIME: 22:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: CARY

DATE/TIME OF TEST: 04/15/2018 23:17:55

---

Diagnostics Check	OK	23:11
Air Blank	0.000	23:12
Control Test	0.080	23:12
Air Blank	0.000	23:13
Subject Sample # 1	0.319	23:13
Breath Volume	1.500	23:13
Air Blank	0.000	23:14
Air Blank	0.000	23:16
Subject Sample # 2	0.325	23:16
Breath Volume	1.550	23:16
Air Blank	0.000	23:17
Control Test	0.080	23:17
Air Blank	0.000	23:17
Diagnostics Check	OK	23:17

---

Data Download Date/Time: 04/29/2018 08:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ANNESE,MORRIS,J  
BREATH TEST OPERATOR AGENCY: West Palm Beach PD  
ARRESTING OFFICER: DILLARD,DANIEL  
ARRESTING OFFICER AGENCY: West Palm Beach PD  
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 152169  
DRY GAS EXPIRY DATE: 10/30/2018  
VIOLATION CODE: DUI  
ARREST TIME: 18:17  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:37

SUBJECT'S LAST NAME: HOOKS

DATE/TIME OF TEST: 04/25/2018 19:11:01

---

Diagnostics Check	OK	19:04
Air Blank	0.000	19:05
Control Test	0.081	19:05
Air Blank	0.000	19:05
Subject Sample # 1	0.000	19:06
Breath Volume	1.921	19:06
Air Blank	0.000	19:07
Air Blank	0.000	19:09
Subject Sample # 2	0.000	19:09
Breath Volume	1.960	19:09
Air Blank	0.000	19:10
Control Test	0.080	19:10
Air Blank	0.000	19:10
Diagnostics Check	OK	19:11

---

Data Download Date/Time: 04/29/2018 08:00



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMAS,MICAH,L  
BREATH TEST OPERATOR AGENCY: West Palm Beach PD  
ARRESTING OFFICER: THOMAS,MICAH  
ARRESTING OFFICER AGENCY: West Palm Beach PD  
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 152169  
DRY GAS EXPIRY DATE: 10/30/2018  
VIOLATION CODE: DUI  
ARREST TIME: 04:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: BRANFIELD

DATE/TIME OF TEST: 04/28/2018 05:21:56

---

Diagnostics Check	OK	05:15
Air Blank	0.000	05:16
Control Test	0.080	05:16
Air Blank	0.000	05:16
Subject Sample # 1	0.259	05:17
Breath Volume	4.496	05:17
Air Blank	0.000	05:18
Air Blank	0.000	05:19
Subject Sample # 2	0.242	05:20
Breath Volume	3.601	05:20
Air Blank	0.000	05:21
Control Test	0.080	05:21
Air Blank	0.000	05:21
Diagnostics Check	OK	05:21

---

Data Download Date/Time: 04/29/2018 08:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMAS,MICAH,L  
BREATH TEST OPERATOR AGENCY: West Palm Beach PD  
ARRESTING OFFICER: THOMAS,MICAH  
ARRESTING OFFICER AGENCY: West Palm Beach PD  
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 152169  
DRY GAS EXPIRY DATE: 10/30/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: SAINT JUSTE

DATE/TIME OF TEST: 04/29/2018 04:08:39

---

Diagnostics Check	OK	03:59
Air Blank	0.000	04:00
Control Test	0.080	04:00
Air Blank	0.000	04:00
Subject Sample # 1	0.220	04:03
Breath Volume	1.535	04:03
Air Blank	0.000	04:04
Air Blank	0.000	04:06
Subject Sample # 2	0.207	04:07
Breath Volume	1.605	04:07
Air Blank	0.000	04:07
Control Test	0.078	04:08
Air Blank	0.000	04:08
Diagnostics Check	OK	04:08

---

Data Download Date/Time: 04/29/2018 08:00

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BURICK,PETER,  
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores  
ARRESTING OFFICER: WALLIS,RICHARD  
ARRESTING OFFICER AGENCY: Holly Hills PD  
LAST AGENCY INSPECTION: 03/04/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 906364  
DRY GAS EXPIRY DATE: 02/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 06:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:09

SUBJECT'S LAST NAME: MARINO

DATE/TIME OF TEST: 04/10/2018 07:39:31

---

Diagnostics Check	OK	07:33
Air Blank	0.000	07:33
Control Test	0.080	07:33
Air Blank	0.000	07:34
Subject Sample # 1	0.171	07:34
Breath Volume	1.773	07:34
Air Blank	0.000	07:35
Air Blank	0.000	07:37
Subject Sample # 2	0.178	07:37
Breath Volume	2.835	07:37
Air Blank	0.000	07:38
Control Test	0.079	07:38
Air Blank	0.000	07:39
Diagnostics Check	OK	07:39

---

Data Download Date/Time: 04/29/2018 02:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOTZ,KATHRYN,N  
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores  
ARRESTING OFFICER: GOTZ,KATHRYN  
ARRESTING OFFICER AGENCY: Daytona Beach Shores  
LAST AGENCY INSPECTION: 03/04/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 906364  
DRY GAS EXPIRY DATE: 02/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 04/16/2018 03:04:59

---

Diagnostics Check		OK	02:51	
Air Blank		0.000	02:52	
Control Test		0.080	02:52	
Air Blank		0.000	02:52	
Subject Sample	# 1	0.206	02:54	
Breath Volume		1.996	02:54	
Air Blank		0.000	02:55	
Air Blank		0.000	02:57	
Subject Sample	# 2	0.153	03:00	Volume Not Met
Breath Volume		0.519	03:00	
Air Blank		0.000	03:00	
Air Blank		0.000	03:02	
Subject Sample	# 3	0.221	03:03	
Breath Volume		1.980	03:03	
Air Blank		0.000	03:04	
Control Test		0.078	03:04	
Air Blank		0.000	03:04	
Diagnostics Check		OK	03:04	

---

Data Download Date/Time: 04/29/2018 02:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOTZ,KATHRYN,N  
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores  
ARRESTING OFFICER: GOTZ,KATHRYN  
ARRESTING OFFICER AGENCY: Daytona Beach Shores  
LAST AGENCY INSPECTION: 03/04/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 906364  
DRY GAS EXPIRY DATE: 02/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: RICHTER

DATE/TIME OF TEST: 04/17/2018 03:44:19

---

Diagnostics Check		OK	03:37
Air Blank		0.000	03:38
Control Test		0.081	03:38
Air Blank		0.000	03:38
Subject Sample	# 1	0.178	03:39
Breath Volume		3.867	03:39
Air Blank		0.000	03:40
Air Blank		0.000	03:41
Subject Sample	# 2	0.167	03:42
Breath Volume		3.285	03:42
Air Blank		0.000	03:43
Control Test		0.080	03:43
Air Blank		0.000	03:44
Diagnostics Check		OK	03:44

---

Data Download Date/Time: 04/29/2018 02:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CASTELLANO,VINCENT,J  
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores  
ARRESTING OFFICER: CASTELLANO,VINCENT  
ARRESTING OFFICER AGENCY: Daytona Beach Shores  
LAST AGENCY INSPECTION: 03/04/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 906364  
DRY GAS EXPIRY DATE: 02/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:39  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: BOYLES

DATE/TIME OF TEST: 04/21/2018 01:14:43

---

Diagnostics Check	OK	01:07
Air Blank	0.000	01:08
Control Test	0.080	01:08
Air Blank	0.000	01:09
Subject Sample # 1	0.062	01:10
Breath Volume	2.085	01:10
Air Blank	0.000	01:10
Air Blank	0.000	01:12
Subject Sample # 2	0.064	01:13
Breath Volume	2.113	01:13
Air Blank	0.000	01:13
Control Test	0.080	01:14
Air Blank	0.000	01:14
Diagnostics Check	OK	01:14

---

Data Download Date/Time: 04/29/2018 02:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOTZ,KATHRYN,N  
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores  
ARRESTING OFFICER: GOTZ,KATHRYN  
ARRESTING OFFICER AGENCY: Daytona Beach Shores  
LAST AGENCY INSPECTION: 03/04/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 906364  
DRY GAS EXPIRY DATE: 02/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 12:38  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: CHASTAIN

DATE/TIME OF TEST: 04/24/2018 01:29:11

---

Diagnostics Check	OK	01:22
Air Blank	0.000	01:23
Control Test	0.079	01:23
Air Blank	0.000	01:24
Subject Sample # 1	0.136	01:24
Breath Volume	1.453	01:24
Air Blank	0.000	01:25
Air Blank	0.000	01:27
Subject Sample # 2	0.139	01:27
Breath Volume	1.554	01:27
Air Blank	0.000	01:28
Control Test	0.079	01:28
Air Blank	0.000	01:29
Diagnostics Check	OK	01:29

---

Data Download Date/Time: 04/29/2018 02:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOTZ,KATHRYN,N  
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores  
ARRESTING OFFICER: GOTZ,KATHRYN  
ARRESTING OFFICER AGENCY: Daytona Beach Shores  
LAST AGENCY INSPECTION: 04/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 906364  
DRY GAS EXPIRY DATE: 02/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:53  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: BELLAMY

DATE/TIME OF TEST: 04/29/2018 03:31:39

---

Diagnostics Check	OK	03:25
Air Blank	0.000	03:25
Control Test	0.079	03:26
Air Blank	0.000	03:26
Subject Sample # 1	0.029	03:27
Breath Volume	1.597	03:27
Air Blank	0.000	03:27
Air Blank	0.000	03:29
Subject Sample # 2	0.026	03:30
Breath Volume	1.175	03:30
Air Blank	0.000	03:30
Control Test	0.079	03:31
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

---

Data Download Date/Time: 05/22/2018 05:24



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RADEMACHER,RICHARD,J  
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores  
ARRESTING OFFICER: RADEMACHER,RICHARD  
ARRESTING OFFICER AGENCY: Daytona Beach Shores  
LAST AGENCY INSPECTION: 04/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 906364  
DRY GAS EXPIRY DATE: 02/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: DISCALA

DATE/TIME OF TEST: 04/29/2018 04:15:34

---

Diagnostics Check	OK	04:06
Air Blank	0.000	04:07
Control Test	0.080	04:07
Air Blank	0.000	04:08
Subject Sample # 1	0.175	04:10
Breath Volume	1.496	04:10
Air Blank	0.000	04:11
Air Blank	0.000	04:13
Subject Sample # 2	0.186	04:14
Breath Volume	1.453	04:14
Air Blank	0.000	04:14
Control Test	0.078	04:14
Air Blank	0.000	04:15
Diagnostics Check	OK	04:15

---

Data Download Date/Time: 05/22/2018 05:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LAWLER,MYLES,J  
BREATH TEST OPERATOR AGENCY: Edgewater PD  
ARRESTING OFFICER: HIRSCH,STEVEN  
ARRESTING OFFICER AGENCY: Edgewater PD  
LAST AGENCY INSPECTION: 03/31/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 765462  
DRY GAS EXPIRY DATE: 04/13/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: ROZIER

DATE/TIME OF TEST: 04/11/2018 01:33:11

---

Diagnostics Check	OK	01:22
Air Blank	0.000	01:23
Control Test	0.078	01:23
Air Blank	0.000	01:24
Subject Sample # 1	0.265	01:26
Breath Volume	1.542	01:26
Air Blank	0.000	01:26
Air Blank	0.000	01:28
Subject Sample # 2	0.267	01:31
Breath Volume	1.335	01:31
Air Blank	0.000	01:32
Control Test	0.077	01:32
Air Blank	0.000	01:33
Diagnostics Check	OK	01:33

---

Data Download Date/Time: 04/27/2018 03:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LAWLER,MYLES,J  
BREATH TEST OPERATOR AGENCY: Edgewater PD  
ARRESTING OFFICER: HIRSCH,STEVEN  
ARRESTING OFFICER AGENCY: Edgewater PD  
LAST AGENCY INSPECTION: 03/31/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 765462  
DRY GAS EXPIRY DATE: 04/13/2020  
VIOLATION CODE: DUI  
ARREST TIME: 19:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: RUSSELL

DATE/TIME OF TEST: 04/13/2018 21:38:12

---

Diagnostics Check	OK	21:31
Air Blank	0.000	21:32
Control Test	0.077	21:32
Air Blank	0.000	21:33
Subject Sample # 1	0.142	21:33
Breath Volume	2.058	21:33
Air Blank	0.000	21:34
Air Blank	0.000	21:36
Subject Sample # 2	0.144	21:36
Breath Volume	1.914	21:36
Air Blank	0.000	21:37
Control Test	0.077	21:37
Air Blank	0.000	21:38
Diagnostics Check	OK	21:38

---

Data Download Date/Time: 04/27/2018 03:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SYDESKI, BLAKE, S  
BREATH TEST OPERATOR AGENCY: Port Orange PD  
ARRESTING OFFICER: WEST, JOHN  
ARRESTING OFFICER AGENCY: FFWCC (Ocala)  
LAST AGENCY INSPECTION: 03/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 609878  
DRY GAS EXPIRY DATE: 11/01/2019  
VIOLATION CODE: BUI  
ARREST TIME: 19:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:19

SUBJECT'S LAST NAME: HUGHES

DATE/TIME OF TEST: 04/03/2018 21:03:56

---

Diagnostics Check	OK	20:56
Air Blank	0.000	20:57
Control Test	0.080	20:57
Air Blank	0.000	20:58
Subject Sample # 1	0.219	20:59
Breath Volume	1.296	20:59
Air Blank	0.000	20:59
Air Blank	0.000	21:01
Subject Sample # 2	0.222	21:02
Breath Volume	1.527	21:02
Air Blank	0.000	21:03
Control Test	0.080	21:03
Air Blank	0.000	21:03
Diagnostics Check	OK	21:03

---

Data Download Date/Time: 04/30/2018 12:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NOLAN,JONATHAN,W  
BREATH TEST OPERATOR AGENCY: Port Orange PD  
ARRESTING OFFICER: MUNOZ,DIEGO  
ARRESTING OFFICER AGENCY: Port Orange PD  
LAST AGENCY INSPECTION: 03/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 609878  
DRY GAS EXPIRY DATE: 11/01/2019  
VIOLATION CODE: DUI  
ARREST TIME: 10:01  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 10:25

SUBJECT'S LAST NAME: BAGNOLI

DATE/TIME OF TEST: 04/05/2018 10:51:48

---

Diagnostics Check	OK	10:45
Air Blank	0.000	10:46
Control Test	0.081	10:46
Air Blank	0.000	10:47
Subject Sample # 1	0.000	10:47
Breath Volume	1.750	10:47
Air Blank	0.000	10:48
Air Blank	0.000	10:50
Subject Sample # 2	0.000	10:50
Breath Volume	2.203	10:50
Air Blank	0.000	10:50
Control Test	0.081	10:51
Air Blank	0.000	10:51
Diagnostics Check	OK	10:51

---

Data Download Date/Time: 04/30/2018 12:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EDWARDS,JUSTIN,D  
BREATH TEST OPERATOR AGENCY: Port Orange PD  
ARRESTING OFFICER: SANDERSON,TAYLOR  
ARRESTING OFFICER AGENCY: South Daytona PD  
LAST AGENCY INSPECTION: 03/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 609878  
DRY GAS EXPIRY DATE: 11/01/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:38  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: MANOULIAN

DATE/TIME OF TEST: 04/06/2018 01:25:32

---

Diagnostics Check	OK	01:24	
Air Blank		01:25	Ambient Fail
Air Blank	0.000	01:25	

---

Data Download Date/Time: 04/30/2018 12:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EDWARDS,JUSTIN,D  
BREATH TEST OPERATOR AGENCY: Port Orange PD  
ARRESTING OFFICER: SANDERSON,TAYLOR  
ARRESTING OFFICER AGENCY: South Daytona PD  
LAST AGENCY INSPECTION: 03/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 609878  
DRY GAS EXPIRY DATE: 11/01/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:38  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: MANOULIAN

DATE/TIME OF TEST: 04/06/2018 01:37:03

---

Diagnostics Check	OK	01:29
Air Blank	0.000	01:29
Control Test	0.080	01:29
Air Blank	0.000	01:30
Subject Sample # 1	0.182	01:32
Breath Volume	1.640	01:32
Air Blank	0.000	01:33
Air Blank	0.000	01:35
Subject Sample # 2	0.178	01:35
Breath Volume	1.906	01:35
Air Blank	0.000	01:36
Control Test	0.079	01:36
Air Blank	0.000	01:36
Diagnostics Check	OK	01:37

---

Data Download Date/Time: 04/30/2018 12:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAINES,CHRISTOPHER,G  
BREATH TEST OPERATOR AGENCY: Port Orange PD  
ARRESTING OFFICER: RAINES,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Port Orange PD  
LAST AGENCY INSPECTION: 03/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 609878  
DRY GAS EXPIRY DATE: 11/01/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:52  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: ELLIS

DATE/TIME OF TEST: 04/14/2018 01:44:18

---

Diagnostics Check		OK	01:38
Air Blank		0.000	01:38
Control Test		0.081	01:38
Air Blank		0.000	01:39
Subject Sample	# 1	0.070	01:39
Breath Volume		1.390	01:39
Air Blank		0.000	01:40
Air Blank		0.000	01:42
Subject Sample	# 2	0.070	01:42
Breath Volume		1.734	01:42
Air Blank		0.000	01:43
Control Test		0.081	01:43
Air Blank		0.000	01:44
Diagnostics Check		OK	01:44

---

Data Download Date/Time: 04/30/2018 12:43



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAINES,CHRISTOPHER,G  
BREATH TEST OPERATOR AGENCY: Port Orange PD  
ARRESTING OFFICER: RAYNOR,JASON  
ARRESTING OFFICER AGENCY: Port Orange PD  
LAST AGENCY INSPECTION: 03/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 609878  
DRY GAS EXPIRY DATE: 11/01/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: RIEHM

DATE/TIME OF TEST: 04/14/2018 02:53:35

---

Diagnostics Check	OK	02:47
Air Blank	0.000	02:47
Control Test	0.081	02:48
Air Blank	0.000	02:48
Subject Sample # 1	0.042	02:49
Breath Volume	4.394	02:49
Air Blank	0.000	02:49
Air Blank	0.000	02:51
Subject Sample # 2	0.044	02:52
Breath Volume	3.117	02:52
Air Blank	0.000	02:52
Control Test	0.082	02:53
Air Blank	0.000	02:53
Diagnostics Check	OK	02:53

---

Data Download Date/Time: 04/30/2018 12:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BENEZETTE,BENJAMIN,J  
BREATH TEST OPERATOR AGENCY: Port Orange PD  
ARRESTING OFFICER: BENEZETTE,BENJAMIN  
ARRESTING OFFICER AGENCY: Port Orange PD  
LAST AGENCY INSPECTION: 03/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 609878  
DRY GAS EXPIRY DATE: 11/01/2019  
VIOLATION CODE: ADMIN  
ARREST TIME: 00:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:14

SUBJECT'S LAST NAME: DAVILA

DATE/TIME OF TEST: 04/17/2018 08:42:16

---

Diagnostics Check	OK	08:36
Air Blank	0.000	08:36
Control Test	0.081	08:37
Air Blank	0.000	08:37
Subject Sample # 1	0.000	08:38
Breath Volume	2.167	08:38
Air Blank	0.000	08:38
Air Blank	0.000	08:40
Subject Sample # 2	0.000	08:40
Breath Volume	1.667	08:40
Air Blank	0.000	08:41
Control Test	0.081	08:41
Air Blank	0.000	08:42
Diagnostics Check	OK	08:42

---

Data Download Date/Time: 04/30/2018 12:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAMARA,CRYSTA,N  
BREATH TEST OPERATOR AGENCY: Port Orange PD  
ARRESTING OFFICER: CAMARA,CRYSTA  
ARRESTING OFFICER AGENCY: Port Orange PD  
LAST AGENCY INSPECTION: 03/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 609878  
DRY GAS EXPIRY DATE: 11/01/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:58  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: TRENT

DATE/TIME OF TEST: 04/19/2018 03:53:37

---

Diagnostics Check	OK	03:47
Air Blank	0.000	03:47
Control Test	0.081	03:48
Air Blank	0.000	03:48
Subject Sample # 1	0.125	03:49
Breath Volume	3.449	03:49
Air Blank	0.000	03:49
Air Blank	0.000	03:51
Subject Sample # 2	0.127	03:52
Breath Volume	3.410	03:52
Air Blank	0.000	03:52
Control Test	0.081	03:53
Air Blank	0.000	03:53
Diagnostics Check	OK	03:53

---

Data Download Date/Time: 04/30/2018 12:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MANDESE,ANGELO,  
BREATH TEST OPERATOR AGENCY: Port Orange PD  
ARRESTING OFFICER: MANDESE,ANGELO  
ARRESTING OFFICER AGENCY: Port Orange PD  
LAST AGENCY INSPECTION: 03/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 609878  
DRY GAS EXPIRY DATE: 11/01/2019  
VIOLATION CODE: ADMIN  
ARREST TIME: 06:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:00

SUBJECT'S LAST NAME: FLEMING

DATE/TIME OF TEST: 04/25/2018 07:41:49

---

Diagnostics Check	OK	07:35
Air Blank	0.000	07:36
Control Test	0.081	07:36
Air Blank	0.000	07:36
Subject Sample # 1	0.000	07:37
Breath Volume	1.828	07:37
Air Blank	0.000	07:37
Air Blank	0.000	07:39
Subject Sample # 2	0.000	07:40
Breath Volume	1.843	07:40
Air Blank	0.000	07:40
Control Test	0.082	07:41
Air Blank	0.000	07:41
Diagnostics Check	OK	07:41

---

Data Download Date/Time: 04/30/2018 12:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EDWARDS,JUSTIN,D  
BREATH TEST OPERATOR AGENCY: Port Orange PD  
ARRESTING OFFICER: EDWARDS,JUSTIN  
ARRESTING OFFICER AGENCY: Port Orange PD  
LAST AGENCY INSPECTION: 03/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 609878  
DRY GAS EXPIRY DATE: 11/01/2019  
VIOLATION CODE: DUI  
ARREST TIME: 04:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: ROZIER

DATE/TIME OF TEST: 04/29/2018 04:57:29

---

Diagnostics Check	OK	04:51
Air Blank	0.000	04:51
Control Test	0.081	04:52
Air Blank	0.000	04:52
Subject Sample # 1	0.000	04:53
Breath Volume	1.558	04:53
Air Blank	0.000	04:53
Air Blank	0.000	04:55
Subject Sample # 2	0.000	04:56
Breath Volume	1.687	04:56
Air Blank	0.000	04:56
Control Test	0.081	04:56
Air Blank	0.000	04:57
Diagnostics Check	OK	04:57

---

Data Download Date/Time: 04/30/2018 12:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001246

INSTRUMENT REGISTERED TO: SOUTH DAYTONA P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHAVES,JOSEPH,A  
BREATH TEST OPERATOR AGENCY: South Daytona PD  
ARRESTING OFFICER: WOODSON,SAMMY  
ARRESTING OFFICER AGENCY: South Daytona PD  
LAST AGENCY INSPECTION: 04/14/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 894572  
DRY GAS EXPIRY DATE: 01/17/2021  
VIOLATION CODE: DUI  
ARREST TIME: 13:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:20

SUBJECT'S LAST NAME: LAWRENCE

DATE/TIME OF TEST: 04/17/2018 14:04:22

---

Diagnostics Check	OK	13:57
Air Blank	0.000	13:58
Control Test	0.082	13:58
Air Blank	0.000	13:58
Subject Sample # 1	0.197	13:59
Breath Volume	1.820	13:59
Air Blank	0.000	13:59
Air Blank	0.000	14:01
Subject Sample # 2	0.195	14:02
Breath Volume	2.078	14:02
Air Blank	0.000	14:03
Control Test	0.082	14:03
Air Blank	0.000	14:04
Diagnostics Check	OK	14:04

---

Data Download Date/Time: 05/17/2018 23:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001246

INSTRUMENT REGISTERED TO: SOUTH DAYTONA P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MELODY,FRANCIS,D  
BREATH TEST OPERATOR AGENCY: South Daytona PD  
ARRESTING OFFICER: GUILLERM,MICHAEL  
ARRESTING OFFICER AGENCY: Port Orange PD  
LAST AGENCY INSPECTION: 04/14/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 894572  
DRY GAS EXPIRY DATE: 01/17/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: DESANTO

DATE/TIME OF TEST: 04/28/2018 04:00:15

---

Diagnostics Check	OK	03:54
Air Blank	0.000	03:54
Control Test	0.082	03:54
Air Blank	0.000	03:55
Subject Sample # 1	0.129	03:55
Breath Volume	1.851	03:55
Air Blank	0.000	03:56
Air Blank	0.000	03:58
Subject Sample # 2	0.125	03:58
Breath Volume	2.007	03:58
Air Blank	0.000	03:59
Control Test	0.082	03:59
Air Blank	0.000	04:00
Diagnostics Check	OK	04:00

---

Data Download Date/Time: 05/17/2018 23:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001246

INSTRUMENT REGISTERED TO: SOUTH DAYTONA P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MELODY,FRANCIS,D  
BREATH TEST OPERATOR AGENCY: South Daytona PD  
ARRESTING OFFICER: FISCHETTI,JAMES  
ARRESTING OFFICER AGENCY: Port Orange PD  
LAST AGENCY INSPECTION: 04/14/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 894572  
DRY GAS EXPIRY DATE: 01/17/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:59  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 04/28/2018 04:12:18

---

Diagnostics Check	OK	04:06
Air Blank	0.000	04:06
Control Test	0.083	04:06
Air Blank	0.000	04:07
Subject Sample # 1	0.170	04:07
Breath Volume	2.066	04:07
Air Blank	0.000	04:08
Air Blank	0.000	04:10
Subject Sample # 2	0.163	04:10
Breath Volume	1.824	04:10
Air Blank	0.000	04:11
Control Test	0.082	04:11
Air Blank	0.000	04:12
Diagnostics Check	OK	04:12

---

Data Download Date/Time: 05/17/2018 23:06



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARRITY, THOMAS, A  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: GARRITY, THOMAS  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 03/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A5  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 23:54  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: CAMARGO

DATE/TIME OF TEST: 04/05/2018 00:45:26

---

Diagnostics Check		OK	00:36
Air Blank		0.000	00:37
Control Test		0.080	00:37
Air Blank		0.000	00:38
Subject Sample	# 1	0.147	00:39
Breath Volume		1.496	00:39
Air Blank		0.000	00:40
Air Blank		0.000	00:42
Subject Sample	# 2	0.135	00:43
Breath Volume		1.351	00:43
Air Blank		0.000	00:44
Control Test		0.080	00:44
Air Blank		0.000	00:45
Diagnostics Check		OK	00:45

---

Data Download Date/Time: 05/15/2018 11:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LAKEMAN,TAYLOR,J  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: ROBIN,GASTON  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 03/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A5  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 00:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: SKORUDE

DATE/TIME OF TEST: 04/07/2018 01:17:01

---

Diagnostics Check	OK	01:12	
Air Blank	0.000	01:12	
Control Test	0.080	01:13	
Air Blank	0.000	01:13	
Subject Sample	# 1	01:15	Subject Test Refused
Breath Volume	0.121	01:15	
Air Blank	0.000	01:16	
Control Test	0.080	01:16	
Air Blank	0.000	01:16	
Diagnostics Check	OK	01:17	

---

Data Download Date/Time: 05/15/2018 11:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LAKEMAN,TAYLOR,J  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: ROBIN,GASTON  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 03/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A5  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 00:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: SKORUDE

DATE/TIME OF TEST: 04/07/2018 01:25:27

---

Diagnostics Check	OK	01:21	
Air Blank	0.000	01:21	
Control Test	0.080	01:22	
Air Blank	0.000	01:22	
Subject Sample	# 1	01:23	Subject Test Refused
Breath Volume	0.058	01:23	
Air Blank	0.000	01:24	
Control Test	0.079	01:24	
Air Blank	0.000	01:25	
Diagnostics Check	OK	01:25	

---

Data Download Date/Time: 05/15/2018 11:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HANNON,LAUREN,F  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: IRIZARRY,HEIDEN  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 03/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A5  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 21:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: CHEWUK

DATE/TIME OF TEST: 04/17/2018 22:06:42

---

Diagnostics Check	OK	22:00
Air Blank	0.000	22:01
Control Test	0.081	22:01
Air Blank	0.000	22:01
Subject Sample # 1	0.299	22:02
Breath Volume	1.496	22:02
Air Blank	0.000	22:03
Air Blank	0.000	22:04
Subject Sample # 2	0.289	22:05
Breath Volume	1.968	22:05
Air Blank	0.000	22:05
Control Test	0.079	22:06
Air Blank	0.000	22:06
Diagnostics Check	OK	22:06

---

Data Download Date/Time: 05/15/2018 11:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARRITY, THOMAS, A  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: GARRITY, THOMAS  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 03/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A5  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 22:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: CRUMPLER

DATE/TIME OF TEST: 04/21/2018 23:19:11

---

Diagnostics Check	OK	23:12
Air Blank	0.000	23:12
Control Test	0.081	23:13
Air Blank	0.000	23:13
Subject Sample # 1	0.217	23:14
Breath Volume	2.179	23:14
Air Blank	0.000	23:15
Air Blank	0.000	23:17
Subject Sample # 2	0.225	23:17
Breath Volume	2.152	23:17
Air Blank	0.000	23:18
Control Test	0.080	23:18
Air Blank	0.000	23:19
Diagnostics Check	OK	23:19

---

Data Download Date/Time: 05/15/2018 11:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMPSON,CALEB,N  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: ROCCHILD,MIGUEL  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 03/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: NIEVES ADORNO

DATE/TIME OF TEST: 04/01/2018 04:45:10

---

Diagnostics Check	OK	04:38
Air Blank	0.000	04:39
Control Test	0.083	04:39
Air Blank	0.000	04:40
Subject Sample # 1	0.171	04:40
Breath Volume	3.207	04:40
Air Blank	0.000	04:41
Air Blank	0.000	04:43
Subject Sample # 2	0.170	04:43
Breath Volume	3.125	04:43
Air Blank	0.000	04:44
Control Test	0.082	04:44
Air Blank	0.000	04:45
Diagnostics Check	OK	04:45

---

Data Download Date/Time: 04/05/2018 15:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARDEN,WILLIAM,L  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: HEBEL,GRIFFIN  
ARRESTING OFFICER AGENCY: Windermere PD  
LAST AGENCY INSPECTION: 03/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:33  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: MALFETANO

DATE/TIME OF TEST: 04/04/2018 02:27:56

---

Diagnostics Check	OK	02:21
Air Blank	0.000	02:21
Control Test	0.081	02:21
Air Blank	0.000	02:22
Subject Sample # 1	0.173	02:23
Breath Volume	2.402	02:23
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample # 2	0.173	02:26
Breath Volume	2.554	02:26
Air Blank	0.000	02:27
Control Test	0.081	02:27
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

---

Data Download Date/Time: 04/05/2018 15:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARDEN,WILLIAM,L  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: COLE,KYLE  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 03/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: TURPIN

DATE/TIME OF TEST: 04/04/2018 04:17:41

---

Diagnostics Check	OK	04:10
Air Blank	0.000	04:11
Control Test	0.082	04:11
Air Blank	0.000	04:12
Subject Sample # 1	0.164	04:12
Breath Volume	1.125	04:12
Air Blank	0.000	04:13
Air Blank	0.000	04:15
Subject Sample # 2	0.160	04:16
Breath Volume	1.488	04:16
Air Blank	0.000	04:16
Control Test	0.082	04:17
Air Blank	0.000	04:17
Diagnostics Check	OK	04:17

---

Data Download Date/Time: 04/05/2018 15:31



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARDEN,WILLIAM,L  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: COOP,JOHN  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 03/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: PENA ARAN

DATE/TIME OF TEST: 04/06/2018 02:36:52

---

Diagnostics Check	OK	02:30
Air Blank	0.000	02:30
Control Test	0.082	02:31
Air Blank	0.000	02:31
Subject Sample # 1	0.140	02:32
Breath Volume	1.207	02:32
Air Blank	0.000	02:33
Air Blank	0.000	02:34
Subject Sample # 2	0.145	02:35
Breath Volume	1.226	02:35
Air Blank	0.000	02:35
Control Test	0.082	02:36
Air Blank	0.000	02:36
Diagnostics Check	OK	02:36

---

Data Download Date/Time: 04/12/2018 11:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARDEN,WILLIAM,L  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: SPEER,ANDREW  
ARRESTING OFFICER AGENCY: Maitland PD  
LAST AGENCY INSPECTION: 03/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:02

SUBJECT'S LAST NAME: UBIERA

DATE/TIME OF TEST: 04/11/2018 05:32:27

---

Diagnostics Check	OK	05:25
Air Blank	0.000	05:25
Control Test	0.083	05:26
Air Blank	0.000	05:26
Subject Sample # 1	0.091	05:27
Breath Volume	1.183	05:27
Air Blank	0.000	05:28
Air Blank	0.000	05:30
Subject Sample # 2	0.097	05:30
Breath Volume	1.496	05:30
Air Blank	0.000	05:31
Control Test	0.081	05:31
Air Blank	0.000	05:32
Diagnostics Check	OK	05:32

---

Data Download Date/Time: 04/12/2018 11:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARDEN,WILLIAM,L  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: KANE,ALEXANDER  
ARRESTING OFFICER AGENCY: Edgewood PD  
LAST AGENCY INSPECTION: 03/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: GLOVER

DATE/TIME OF TEST: 04/12/2018 04:14:52

---

Diagnostics Check	OK	04:12	
Air Blank	0.000	04:12	
Control Test	0.082	04:12	
Air Blank	0.000	04:13	
Subject Sample	# 1	04:13	Subject Test Refused
Breath Volume	0.000	04:13	
Air Blank	0.000	04:13	
Control Test	0.081	04:14	
Air Blank	0.000	04:14	
Diagnostics Check	OK	04:14	

---

Data Download Date/Time: 04/12/2018 11:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARDEN,WILLIAM,L  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: HAWKINS,DEBBIE  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 03/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: HARPER

DATE/TIME OF TEST: 04/13/2018 02:54:00

---

Diagnostics Check	OK	02:47
Air Blank	0.000	02:47
Control Test	0.083	02:48
Air Blank	0.000	02:48
Subject Sample # 1	0.000	02:49
Breath Volume	3.011	02:49
Air Blank	0.000	02:50
Air Blank	0.000	02:52
Subject Sample # 2	0.000	02:52
Breath Volume	2.238	02:52
Air Blank	0.000	02:53
Control Test	0.080	02:53
Air Blank	0.000	02:53
Diagnostics Check	OK	02:54

---

Data Download Date/Time: 04/17/2018 12:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARDEN,WILLIAM,L  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: PARADIS,GEORGE  
ARRESTING OFFICER AGENCY: Maitland PD  
LAST AGENCY INSPECTION: 03/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: STEELE III

DATE/TIME OF TEST: 04/17/2018 02:04:55

---

Diagnostics Check	OK	01:58
Air Blank	0.000	01:58
Control Test	0.083	01:59
Air Blank	0.000	01:59
Subject Sample # 1	0.000	02:00
Breath Volume	3.589	02:00
Air Blank	0.000	02:01
Air Blank	0.000	02:03
Subject Sample # 2	0.000	02:03
Breath Volume	4.730	02:03
Air Blank	0.000	02:03
Control Test	0.081	02:04
Air Blank	0.000	02:04
Diagnostics Check	OK	02:04

---

Data Download Date/Time: 04/17/2018 12:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARDEN,WILLIAM,L  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: MORALES,JOSE  
ARRESTING OFFICER AGENCY: Central FL Univ PD  
LAST AGENCY INSPECTION: 03/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: PRIEST

DATE/TIME OF TEST: 04/17/2018 03:46:09

---

Diagnostics Check	OK	03:43	
Air Blank	0.000	03:43	
Control Test	0.082	03:44	
Air Blank	0.000	03:44	
Subject Sample	# 1	03:44	Subject Test Refused
Breath Volume	0.000	03:44	
Air Blank	0.000	03:45	
Control Test	0.082	03:45	
Air Blank	0.000	03:46	
Diagnostics Check	OK	03:46	

---

Data Download Date/Time: 04/17/2018 12:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ELBON,NICHOLAS,C  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: GALLUP,ELIAS  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 04/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 18:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:29

SUBJECT'S LAST NAME: KLINE

DATE/TIME OF TEST: 04/19/2018 19:58:30

---

Diagnostics Check	OK	19:51
Air Blank	0.000	19:52
Control Test	0.082	19:52
Air Blank	0.000	19:53
Subject Sample # 1	0.243	19:53
Breath Volume	2.046	19:53
Air Blank	0.000	19:54
Air Blank	0.000	19:56
Subject Sample # 2	0.247	19:56
Breath Volume	2.046	19:56
Air Blank	0.000	19:57
Control Test	0.081	19:57
Air Blank	0.000	19:58
Diagnostics Check	OK	19:58

---

Data Download Date/Time: 04/26/2018 14:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ELBON,NICHOLAS,C  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: RIVERA,VICTER  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 04/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: BURGOS CORREA

DATE/TIME OF TEST: 04/20/2018 03:06:47

---

Diagnostics Check	OK	02:59
Air Blank	0.000	02:59
Control Test	0.082	02:59
Air Blank	0.000	03:00
Subject Sample # 1	0.132	03:01
Breath Volume	1.714	03:01
Air Blank	0.000	03:02
Air Blank	0.000	03:04
Subject Sample # 2	0.128	03:05
Breath Volume	2.523	03:05
Air Blank	0.000	03:05
Control Test	0.079	03:06
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

---

Data Download Date/Time: 04/26/2018 14:20



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: SAMMELIAN,ALBERT  
ARRESTING OFFICER AGENCY: Central FL Univ PD  
LAST AGENCY INSPECTION: 04/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: ISLAM

DATE/TIME OF TEST: 04/20/2018 23:51:12

---

Diagnostics Check	OK	23:44
Air Blank	0.000	23:44
Control Test	0.081	23:45
Air Blank	0.000	23:45
Subject Sample # 1	0.088	23:46
Breath Volume	1.250	23:46
Air Blank	0.000	23:47
Air Blank	0.000	23:49
Subject Sample # 2	0.092	23:49
Breath Volume	1.726	23:49
Air Blank	0.000	23:50
Control Test	0.081	23:50
Air Blank	0.000	23:51
Diagnostics Check	OK	23:51

---

Data Download Date/Time: 04/26/2018 14:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ELBON,NICHOLAS,C  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: AZEVEDO,FABIO  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 04/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: PENA

DATE/TIME OF TEST: 04/27/2018 04:34:50

---

Diagnostics Check	OK	04:27
Air Blank	0.000	04:27
Control Test	0.082	04:28
Air Blank	0.000	04:28
Subject Sample # 1	0.173	04:29
Breath Volume	1.593	04:29
Air Blank	0.000	04:30
Air Blank	0.000	04:32
Subject Sample # 2	0.186	04:33
Breath Volume	3.519	04:33
Air Blank	0.000	04:33
Control Test	0.080	04:34
Air Blank	0.000	04:34
Diagnostics Check	OK	04:34

---

Data Download Date/Time: 05/03/2018 16:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ELBON,NICHOLAS,C  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: PUPO,ANDREW  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 04/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: RAYSOR

DATE/TIME OF TEST: 04/29/2018 00:11:14

---

Diagnostics Check		OK	00:02
Air Blank		0.000	00:03
Control Test		0.082	00:03
Air Blank		0.000	00:03
Subject Sample	# 1	0.254	00:05
Breath Volume		1.335	00:05
Air Blank		0.000	00:06
Air Blank		0.000	00:08
Subject Sample	# 2	0.248	00:09
Breath Volume		1.195	00:09
Air Blank		0.000	00:10
Control Test		0.079	00:10
Air Blank		0.000	00:11
Diagnostics Check		OK	00:11

---

Data Download Date/Time: 05/03/2018 16:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ELBON,NICHOLAS,C  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: MAYFIELD,JOHN  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 04/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:35  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: WARREN

DATE/TIME OF TEST: 04/29/2018 03:29:36

---

Diagnostics Check	OK	03:22
Air Blank	0.000	03:23
Control Test	0.082	03:23
Air Blank	0.000	03:24
Subject Sample # 1	0.184	03:24
Breath Volume	2.429	03:24
Air Blank	0.000	03:25
Air Blank	0.000	03:27
Subject Sample # 2	0.183	03:27
Breath Volume	1.890	03:27
Air Blank	0.000	03:28
Control Test	0.081	03:28
Air Blank	0.000	03:29
Diagnostics Check	OK	03:29

---

Data Download Date/Time: 05/03/2018 16:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TIAMSON,DAWN,P  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: VAUGHN,JACOB  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 03/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 04/01/2018 02:30:50

---

Diagnostics Check		OK	02:24
Air Blank		0.000	02:25
Control Test		0.081	02:25
Air Blank		0.000	02:26
Subject Sample	# 1	0.134	02:26
Breath Volume		2.687	02:26
Air Blank		0.000	02:27
Air Blank		0.000	02:29
Subject Sample	# 2	0.135	02:29
Breath Volume		2.605	02:29
Air Blank		0.000	02:29
Control Test		0.081	02:30
Air Blank		0.000	02:30
Diagnostics Check		OK	02:30

---

Data Download Date/Time: 04/05/2018 15:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DENSON, TRAVIS, W  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: CAMPBELL, CRAIG  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 03/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: NUNEZ

DATE/TIME OF TEST: 04/01/2018 04:34:50

---

Diagnostics Check	OK	04:26
Air Blank	0.000	04:27
Control Test	0.080	04:27
Air Blank	0.000	04:28
Subject Sample # 1	0.133	04:30
Breath Volume	1.160	04:30
Air Blank	0.000	04:30
Air Blank	0.000	04:32
Subject Sample # 2	0.128	04:33
Breath Volume	1.492	04:33
Air Blank	0.000	04:33
Control Test	0.080	04:34
Air Blank	0.000	04:34
Diagnostics Check	OK	04:34

---

Data Download Date/Time: 04/05/2018 15:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: WILKES,DAVID  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 03/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 04:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:33

SUBJECT'S LAST NAME: INTHAVONGSA

DATE/TIME OF TEST: 04/01/2018 05:02:36

---

Diagnostics Check	OK	04:56
Air Blank	0.000	04:56
Control Test	0.080	04:57
Air Blank	0.000	04:57
Subject Sample # 1	0.124	04:58
Breath Volume	3.085	04:58
Air Blank	0.000	04:58
Air Blank	0.000	05:00
Subject Sample # 2	0.121	05:01
Breath Volume	3.011	05:01
Air Blank	0.000	05:01
Control Test	0.080	05:02
Air Blank	0.000	05:02
Diagnostics Check	OK	05:02

---

Data Download Date/Time: 04/05/2018 15:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TIAMSON,DAWN,P  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: CORNWELL,BRANDON  
ARRESTING OFFICER AGENCY: OTHER Local  
LAST AGENCY INSPECTION: 03/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:48  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: SOTO

DATE/TIME OF TEST: 04/02/2018 03:31:41

---

Diagnostics Check		OK	03:25
Air Blank		0.000	03:26
Control Test		0.080	03:26
Air Blank		0.000	03:27
Subject Sample	# 1	0.208	03:27
Breath Volume		1.300	03:27
Air Blank		0.000	03:28
Air Blank		0.000	03:29
Subject Sample	# 2	0.204	03:30
Breath Volume		2.160	03:30
Air Blank		0.000	03:30
Control Test		0.080	03:31
Air Blank		0.000	03:31
Diagnostics Check		OK	03:31

---

Data Download Date/Time: 04/05/2018 15:34



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMPSON,CALEB,N  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: BAKER,ANTHONY  
ARRESTING OFFICER AGENCY: FHP Troop K  
LAST AGENCY INSPECTION: 03/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: YUHENG

DATE/TIME OF TEST: 04/02/2018 05:20:44

---

Diagnostics Check	OK	05:13
Air Blank	0.000	05:13
Control Test	0.081	05:14
Air Blank	0.000	05:14
Subject Sample # 1	0.153	05:15
Breath Volume	3.757	05:15
Air Blank	0.000	05:16
Air Blank	0.000	05:18
Subject Sample # 2	0.157	05:19
Breath Volume	4.031	05:19
Air Blank	0.000	05:19
Control Test	0.080	05:20
Air Blank	0.000	05:20
Diagnostics Check	OK	05:20

---

Data Download Date/Time: 04/05/2018 15:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TIAMSON,DAWN,P  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: MORALES,JOSE  
ARRESTING OFFICER AGENCY: Central FL Univ PD  
LAST AGENCY INSPECTION: 03/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: FOSKETT

DATE/TIME OF TEST: 04/03/2018 03:57:49

---

Diagnostics Check	OK	03:51
Air Blank	0.000	03:52
Control Test	0.080	03:52
Air Blank	0.000	03:52
Subject Sample # 1	0.128	03:53
Breath Volume	3.203	03:53
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample # 2	0.128	03:56
Breath Volume	3.105	03:56
Air Blank	0.000	03:56
Control Test	0.080	03:57
Air Blank	0.000	03:57
Diagnostics Check	OK	03:57

---

Data Download Date/Time: 04/05/2018 15:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMPSON,CALEB,N  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: ROBBINS,NATHAN  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 03/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:01  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:08

SUBJECT'S LAST NAME: EVERT

DATE/TIME OF TEST: 04/03/2018 21:38:40

---

Diagnostics Check	OK	21:32
Air Blank	0.000	21:32
Control Test	0.081	21:33
Air Blank	0.000	21:33
Subject Sample # 1	0.234	21:34
Breath Volume	1.566	21:34
Air Blank	0.000	21:34
Air Blank	0.000	21:36
Subject Sample # 2	0.239	21:37
Breath Volume	2.750	21:37
Air Blank	0.000	21:37
Control Test	0.081	21:38
Air Blank	0.000	21:38
Diagnostics Check	OK	21:38

---

Data Download Date/Time: 04/05/2018 15:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMPSON,CALEB,N  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: RIVERA,VICTOR  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 03/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: GONZALEZMONDONALDO

DATE/TIME OF TEST: 04/04/2018 04:00:05

---

Diagnostics Check	OK	03:57	
Air Blank	0.000	03:57	
Control Test	0.081	03:58	
Air Blank	0.000	03:58	
Subject Sample	# 1	03:58	Subject Test Refused
Breath Volume	0.000	03:58	
Air Blank	0.000	03:59	
Control Test	0.080	03:59	
Air Blank	0.000	04:00	
Diagnostics Check	OK	04:00	

---

Data Download Date/Time: 04/05/2018 15:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ELBON,NICHOLAS,C  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: SHEIMAN,RICHARD  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 03/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: FLYNN

DATE/TIME OF TEST: 04/05/2018 00:07:33

---

Diagnostics Check		OK	00:01
Air Blank		0.000	00:01
Control Test		0.080	00:02
Air Blank		0.000	00:02
Subject Sample	# 1	0.101	00:03
Breath Volume		2.242	00:03
Air Blank		0.000	00:03
Air Blank		0.000	00:05
Subject Sample	# 2	0.104	00:06
Breath Volume		2.781	00:06
Air Blank		0.000	00:06
Control Test		0.080	00:06
Air Blank		0.000	00:07
Diagnostics Check		OK	00:07

---

Data Download Date/Time: 04/05/2018 15:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ELBON,NICHOLAS,C  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: HERNANDEZ,EDUARDO  
ARRESTING OFFICER AGENCY: Ocoee PD  
LAST AGENCY INSPECTION: 03/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: MARAJ

DATE/TIME OF TEST: 04/06/2018 01:50:10

---

Diagnostics Check	OK	01:42
Air Blank	0.000	01:42
Control Test	0.080	01:42
Air Blank	0.000	01:43
Subject Sample # 1	0.272	01:43
Breath Volume	1.347	01:43
Air Blank	0.000	01:44
Air Blank	0.000	01:46
Subject Sample # 2	0.254	01:48
Breath Volume	1.664	01:48
Air Blank	0.000	01:49
Control Test	0.080	01:49
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

---

Data Download Date/Time: 04/12/2018 11:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TIAMSON,DAWN,P  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: MENDOZA,ANGEL  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 03/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: TORRESSANCHEZ

DATE/TIME OF TEST: 04/06/2018 22:02:43

---

Diagnostics Check	OK	21:55
Air Blank	0.000	21:55
Control Test	0.081	21:55
Air Blank	0.000	21:56
Subject Sample # 1	0.203	21:57
Breath Volume	1.433	21:57
Air Blank	0.000	21:58
Air Blank	0.000	22:00
Subject Sample # 2	0.217	22:01
Breath Volume	2.601	22:01
Air Blank	0.000	22:01
Control Test	0.080	22:02
Air Blank	0.000	22:02
Diagnostics Check	OK	22:02

---

Data Download Date/Time: 04/12/2018 11:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TIAMSON,DAWN,P  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: SPENCER,JOHN  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 03/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: FLYNN

DATE/TIME OF TEST: 04/07/2018 02:30:47

---

Diagnostics Check	OK	02:23
Air Blank	0.000	02:24
Control Test	0.081	02:24
Air Blank	0.000	02:24
Subject Sample # 1	0.000	02:25
Breath Volume	1.175	02:25
Air Blank	0.000	02:25
Air Blank	0.000	02:27
Subject Sample # 2	0.000	02:29
Breath Volume	1.488	02:29
Air Blank	0.000	02:29
Control Test	0.081	02:30
Air Blank	0.000	02:30
Diagnostics Check	OK	02:30

---

Data Download Date/Time: 04/12/2018 11:40



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TIAMSON,DAWN,P  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: EBY,ELIZABETH  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 03/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: PIERRE

DATE/TIME OF TEST: 04/08/2018 02:50:45

---

Diagnostics Check		OK	02:39	
Air Blank		0.000	02:39	
Control Test		0.080	02:40	
Air Blank		0.000	02:40	
Subject Sample	# 1	0.115	02:43	Volume Not Met
Breath Volume		0.000	02:43	
Air Blank		0.000	02:44	
Air Blank		0.000	02:46	
Subject Sample	# 2	0.136	02:49	Volume Not Met
Breath Volume		0.015	02:49	
Air Blank		0.000	02:49	
Control Test		0.080	02:50	
Air Blank		0.000	02:50	
Diagnostics Check		OK	02:50	

---

Data Download Date/Time: 04/12/2018 11:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVIS,COURTNEY,L  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: CRUZ,ALEXANDER  
ARRESTING OFFICER AGENCY: Maitland PD  
LAST AGENCY INSPECTION: 03/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 04:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:23

SUBJECT'S LAST NAME: MALDONADO

DATE/TIME OF TEST: 04/08/2018 05:51:08

---

Diagnostics Check		OK	05:44
Air Blank		0.000	05:45
Control Test		0.080	05:45
Air Blank		0.000	05:46
Subject Sample	# 1	0.188	05:46
Breath Volume		2.566	05:46
Air Blank		0.000	05:47
Air Blank		0.000	05:49
Subject Sample	# 2	0.181	05:49
Breath Volume		2.113	05:49
Air Blank		0.000	05:50
Control Test		0.080	05:50
Air Blank		0.000	05:51
Diagnostics Check		OK	05:51

---

Data Download Date/Time: 04/12/2018 11:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMPSON,CALEB,N  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: MILLS,RESHARD  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 03/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 05:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:03

SUBJECT'S LAST NAME: GILLENS

DATE/TIME OF TEST: 04/08/2018 06:33:55

---

Diagnostics Check	OK	06:26
Air Blank	0.000	06:26
Control Test	0.080	06:27
Air Blank	0.000	06:27
Subject Sample # 1	0.191	06:29
Breath Volume	1.437	06:29
Air Blank	0.000	06:29
Air Blank	0.000	06:31
Subject Sample # 2	0.196	06:32
Breath Volume	1.125	06:32
Air Blank	0.000	06:33
Control Test	0.079	06:33
Air Blank	0.000	06:33
Diagnostics Check	OK	06:33

---

Data Download Date/Time: 04/12/2018 11:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TIAMSON,DAWN,P  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: HAKMAN,JASON  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 03/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 04/09/2018 00:30:11

---

Diagnostics Check		OK	00:20
Air Blank		0.000	00:21
Control Test		0.082	00:21
Air Blank		0.000	00:22
Subject Sample	# 1	0.218	00:25
Breath Volume		1.777	00:25
Air Blank		0.000	00:26
Air Blank		0.000	00:27
Subject Sample	# 2	0.207	00:28
Breath Volume		1.195	00:28
Air Blank		0.000	00:29
Control Test		0.080	00:29
Air Blank		0.000	00:30
Diagnostics Check		OK	00:30

---

Data Download Date/Time: 04/12/2018 11:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TIAMSON,DAWN,P  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: SORRELLS,BRIAN  
ARRESTING OFFICER AGENCY: Apopka PD  
LAST AGENCY INSPECTION: 03/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 16:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:35

SUBJECT'S LAST NAME: PACE

DATE/TIME OF TEST: 04/09/2018 18:11:27

---

Diagnostics Check		OK	18:00	
Air Blank		0.000	18:00	
Control Test		0.080	18:00	
Air Blank		0.000	18:01	
Subject Sample	# 1	0.035	18:04	Volume Not Met
Breath Volume		0.308	18:04	
Air Blank		0.000	18:04	
Air Blank		0.000	18:06	
Subject Sample	# 2	0.036	18:10	Volume Not Met
Breath Volume		0.402	18:10	
Air Blank		0.000	18:10	
Control Test		0.081	18:10	
Air Blank		0.000	18:11	
Diagnostics Check		OK	18:11	

---

Data Download Date/Time: 04/12/2018 11:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TIAMSON,DAWN,P  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: SORRELLS,BRIAN  
ARRESTING OFFICER AGENCY: Apopka PD  
LAST AGENCY INSPECTION: 03/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 16:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:35

SUBJECT'S LAST NAME: PACE

DATE/TIME OF TEST: 04/09/2018 18:19:23

---

Diagnostics Check	OK	18:13
Air Blank	0.000	18:14
Control Test	0.081	18:14
Air Blank	0.000	18:14
Subject Sample # 1	0.034	18:15
Breath Volume	1.175	18:15
Air Blank	0.000	18:15
Air Blank	0.000	18:17
Subject Sample # 2	0.032	18:17
Breath Volume	1.105	18:17
Air Blank	0.000	18:18
Control Test	0.080	18:18
Air Blank	0.000	18:19
Diagnostics Check	OK	18:19

---

Data Download Date/Time: 04/12/2018 11:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ELBON,NICHOLAS,C  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: VENTURA,CHRISTOPHER  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 03/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:31  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:01

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 04/09/2018 23:32:19

---

Diagnostics Check	OK	23:25
Air Blank	0.000	23:25
Control Test	0.080	23:25
Air Blank	0.000	23:26
Subject Sample # 1	0.391	23:27
Breath Volume	2.250	23:27
Air Blank	0.000	23:27
Air Blank	0.000	23:29
Subject Sample # 2	0.372	23:30
Breath Volume	2.269	23:30
Air Blank	0.000	23:31
Control Test	0.078	23:31
Air Blank	0.000	23:32
Diagnostics Check	OK	23:32

---

Data Download Date/Time: 04/12/2018 11:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMPSON,CALEB,N  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: RODRIGUEZ,YESSICA  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 03/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: MCCOY

DATE/TIME OF TEST: 04/11/2018 01:23:43

---

Diagnostics Check	OK	01:17
Air Blank	0.000	01:17
Control Test	0.081	01:17
Air Blank	0.000	01:18
Subject Sample # 1	0.140	01:19
Breath Volume	2.621	01:19
Air Blank	0.000	01:19
Air Blank	0.000	01:21
Subject Sample # 2	0.141	01:22
Breath Volume	2.324	01:22
Air Blank	0.000	01:22
Control Test	0.081	01:23
Air Blank	0.000	01:23
Diagnostics Check	OK	01:23

---

Data Download Date/Time: 04/12/2018 11:40



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: CAMPBELL,CRAIG  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 03/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: GUERRERO

DATE/TIME OF TEST: 04/12/2018 03:28:08

---

Diagnostics Check	OK	03:21
Air Blank	0.000	03:22
Control Test	0.081	03:22
Air Blank	0.000	03:23
Subject Sample # 1	0.187	03:23
Breath Volume	1.437	03:23
Air Blank	0.000	03:24
Air Blank	0.000	03:26
Subject Sample # 2	0.182	03:26
Breath Volume	1.949	03:26
Air Blank	0.000	03:27
Control Test	0.080	03:27
Air Blank	0.000	03:28
Diagnostics Check	OK	03:28

---

Data Download Date/Time: 04/12/2018 11:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: RIVERA,VICTOR  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 03/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 04:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 04/12/2018 05:16:25

---

Diagnostics Check		OK	05:10
Air Blank		0.000	05:10
Control Test		0.081	05:11
Air Blank		0.000	05:11
Subject Sample	# 1	0.103	05:12
Breath Volume		2.765	05:12
Air Blank		0.000	05:12
Air Blank		0.000	05:14
Subject Sample	# 2	0.106	05:14
Breath Volume		2.777	05:14
Air Blank		0.000	05:15
Control Test		0.081	05:15
Air Blank		0.000	05:16
Diagnostics Check		OK	05:16

---

Data Download Date/Time: 04/12/2018 11:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: VAUGHN,JACOB  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 03/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: SKELIN

DATE/TIME OF TEST: 04/13/2018 02:20:20

---

Diagnostics Check	OK	02:11
Air Blank	0.000	02:12
Control Test	0.080	02:12
Air Blank	0.000	02:13
Subject Sample # 1	0.100	02:14
Breath Volume	1.757	02:14
Air Blank	0.000	02:15
Air Blank	0.000	02:17
Subject Sample # 2	0.101	02:18
Breath Volume	1.128	02:18
Air Blank	0.000	02:19
Control Test	0.081	02:19
Air Blank	0.000	02:20
Diagnostics Check	OK	02:20

---

Data Download Date/Time: 04/17/2018 11:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DENSON, TRAVIS, W  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: WEAVER, WILLIAM  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 03/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:08  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: ERLBACHER

DATE/TIME OF TEST: 04/13/2018 02:49:31

---

Diagnostics Check	OK	02:43
Air Blank	0.000	02:43
Control Test	0.081	02:44
Air Blank	0.000	02:44
Subject Sample # 1	0.172	02:45
Breath Volume	2.851	02:45
Air Blank	0.000	02:45
Air Blank	0.000	02:47
Subject Sample # 2	0.180	02:48
Breath Volume	2.703	02:48
Air Blank	0.000	02:48
Control Test	0.080	02:48
Air Blank	0.000	02:49
Diagnostics Check	OK	02:49

---

Data Download Date/Time: 04/17/2018 11:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DENSON, TRAVIS, W  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: WEAVER, WILLIAM  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 03/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 19:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: AYALA ALICEA

DATE/TIME OF TEST: 04/13/2018 20:01:49

---

Diagnostics Check	OK	19:55
Air Blank	0.000	19:56
Control Test	0.080	19:56
Air Blank	0.000	19:57
Subject Sample # 1	0.383	19:57
Breath Volume	1.101	19:57
Air Blank	0.000	19:58
Air Blank	0.000	19:59
Subject Sample # 2	0.386	20:00
Breath Volume	1.273	20:00
Air Blank	0.000	20:00
Control Test	0.080	20:01
Air Blank	0.000	20:01
Diagnostics Check	OK	20:01

---

Data Download Date/Time: 04/17/2018 11:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: HUTTO,ROSS  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 03/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: ROSARIO JR

DATE/TIME OF TEST: 04/14/2018 00:15:19

---

Diagnostics Check		OK	00:09
Air Blank		0.000	00:09
Control Test		0.080	00:09
Air Blank		0.000	00:10
Subject Sample	# 1	0.084	00:10
Breath Volume		2.640	00:10
Air Blank		0.000	00:11
Air Blank		0.000	00:13
Subject Sample	# 2	0.087	00:13
Breath Volume		1.214	00:13
Air Blank		0.000	00:14
Control Test		0.081	00:14
Air Blank		0.000	00:15
Diagnostics Check		OK	00:15

---

Data Download Date/Time: 04/17/2018 11:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: ROBERTSON,DANIEL  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 03/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: GAILLARD

DATE/TIME OF TEST: 04/14/2018 04:22:58

---

Diagnostics Check	OK	04:13
Air Blank	0.000	04:14
Control Test	0.080	04:14
Air Blank	0.000	04:14
Subject Sample # 1	0.143	04:17
Breath Volume	1.996	04:17
Air Blank	0.000	04:17
Air Blank	0.000	04:19
Subject Sample # 2	0.129	04:21
Breath Volume	1.621	04:21
Air Blank	0.000	04:22
Control Test	0.080	04:22
Air Blank	0.000	04:22
Diagnostics Check	OK	04:22

---

Data Download Date/Time: 04/17/2018 11:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ, MICHAEL, J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: GAZZO, JOSEPH  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 03/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 13:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:20

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 04/14/2018 14:52:45

---

Diagnostics Check	OK	14:46
Air Blank	0.000	14:46
Control Test	0.081	14:46
Air Blank	0.000	14:47
Subject Sample # 1	0.256	14:48
Breath Volume	2.074	14:48
Air Blank	0.000	14:48
Air Blank	0.000	14:50
Subject Sample # 2	0.247	14:51
Breath Volume	1.378	14:51
Air Blank	0.000	14:51
Control Test	0.080	14:52
Air Blank	0.000	14:52
Diagnostics Check	OK	14:52

---

Data Download Date/Time: 04/17/2018 11:20



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: GIARRATANA,JOHN  
ARRESTING OFFICER AGENCY: Apopka PD  
LAST AGENCY INSPECTION: 03/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: CHARNECO

DATE/TIME OF TEST: 04/15/2018 02:45:38

---

Diagnostics Check	OK	02:42	
Air Blank	0.000	02:42	
Control Test	0.081	02:43	
Air Blank	0.000	02:43	
Subject Sample	# 1	02:44	Interferent Detect
Breath Volume	1.757	02:44	
Air Blank	0.000	02:44	
Control Test	0.081	02:45	
Air Blank	0.000	02:45	
Diagnostics Check	OK	02:45	

---

Data Download Date/Time: 04/17/2018 11:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMPSON,CALEB,N  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: BECTON,WILLIAM  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 03/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: ODDO

DATE/TIME OF TEST: 04/15/2018 04:18:10

---

Diagnostics Check	OK	04:11
Air Blank	0.000	04:12
Control Test	0.081	04:12
Air Blank	0.000	04:12
Subject Sample # 1	0.135	04:13
Breath Volume	3.980	04:13
Air Blank	0.000	04:14
Air Blank	0.000	04:16
Subject Sample # 2	0.134	04:16
Breath Volume	4.074	04:16
Air Blank	0.000	04:17
Control Test	0.080	04:17
Air Blank	0.000	04:18
Diagnostics Check	OK	04:18

---

Data Download Date/Time: 04/17/2018 11:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ELBON,NICHOLAS,C  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: DOWSON,AMANDA  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 03/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 04/16/2018 01:03:45

---

Diagnostics Check	OK	00:57
Air Blank	0.000	00:58
Control Test	0.082	00:58
Air Blank	0.000	00:59
Subject Sample # 1	0.196	00:59
Breath Volume	2.894	00:59
Air Blank	0.000	00:59
Air Blank	0.000	01:01
Subject Sample # 2	0.205	01:02
Breath Volume	3.464	01:02
Air Blank	0.000	01:02
Control Test	0.081	01:03
Air Blank	0.000	01:03
Diagnostics Check	OK	01:03

---

Data Download Date/Time: 04/17/2018 11:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMPSON,CALEB,N  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: ROTARIUS,JOSHUA  
ARRESTING OFFICER AGENCY: Maitland PD  
LAST AGENCY INSPECTION: 03/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: DALEY

DATE/TIME OF TEST: 04/16/2018 03:26:25

---

Diagnostics Check		OK	03:18
Air Blank		0.000	03:19
Control Test		0.080	03:19
Air Blank		0.000	03:20
Subject Sample	# 1	0.054	03:21
Breath Volume		1.792	03:21
Air Blank		0.000	03:22
Air Blank		0.000	03:24
Subject Sample	# 2	0.056	03:24
Breath Volume		1.425	03:24
Air Blank		0.000	03:25
Control Test		0.081	03:25
Air Blank		0.000	03:26
Diagnostics Check		OK	03:26

---

Data Download Date/Time: 04/17/2018 11:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ELBON,NICHOLAS,C  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: BROWN,DARRELL  
ARRESTING OFFICER AGENCY: Winter Garden PD  
LAST AGENCY INSPECTION: 03/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 19:01  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:14

SUBJECT'S LAST NAME: MANN

DATE/TIME OF TEST: 04/16/2018 20:43:56

---

Diagnostics Check	OK	20:36
Air Blank	0.000	20:37
Control Test	0.080	20:37
Air Blank	0.000	20:38
Subject Sample # 1	0.224	20:39
Breath Volume	2.027	20:39
Air Blank	0.000	20:39
Air Blank	0.000	20:41
Subject Sample # 2	0.225	20:42
Breath Volume	2.914	20:42
Air Blank	0.000	20:43
Control Test	0.080	20:43
Air Blank	0.000	20:43
Diagnostics Check	OK	20:43

---

Data Download Date/Time: 04/17/2018 11:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMPSON,CALEB,N  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: NEWELL,JEFF  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 03/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: RODARTE

DATE/TIME OF TEST: 04/17/2018 02:15:44

---

Diagnostics Check		OK	02:09
Air Blank		0.000	02:09
Control Test		0.081	02:09
Air Blank		0.000	02:10
Subject Sample	# 1	0.150	02:11
Breath Volume		1.957	02:11
Air Blank		0.000	02:11
Air Blank		0.000	02:13
Subject Sample	# 2	0.146	02:14
Breath Volume		1.957	02:14
Air Blank		0.000	02:14
Control Test		0.080	02:15
Air Blank		0.000	02:15
Diagnostics Check		OK	02:15

---

Data Download Date/Time: 04/17/2018 11:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: LANDIN, CARLOS  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 04/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: MIDDLEBROOKS

DATE/TIME OF TEST: 04/18/2018 01:55:40

---

Diagnostics Check	OK	01:48
Air Blank	0.000	01:48
Control Test	0.081	01:48
Air Blank	0.000	01:49
Subject Sample # 1	0.159	01:50
Breath Volume	1.167	01:50
Air Blank	0.000	01:50
Air Blank	0.000	01:52
Subject Sample # 2	0.159	01:54
Breath Volume	1.335	01:54
Air Blank	0.000	01:54
Control Test	0.080	01:55
Air Blank	0.000	01:55
Diagnostics Check	OK	01:55

---

Data Download Date/Time: 04/26/2018 14:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: CAMPBELL, CRAIG  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 04/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: SILBANUZ

DATE/TIME OF TEST: 04/18/2018 03:36:16

---

Diagnostics Check	OK	03:30
Air Blank	0.000	03:30
Control Test	0.080	03:30
Air Blank	0.000	03:31
Subject Sample # 1	0.212	03:31
Breath Volume	3.425	03:31
Air Blank	0.000	03:32
Air Blank	0.000	03:34
Subject Sample # 2	0.214	03:34
Breath Volume	3.292	03:34
Air Blank	0.000	03:35
Control Test	0.080	03:35
Air Blank	0.000	03:36
Diagnostics Check	OK	03:36

---

Data Download Date/Time: 04/26/2018 14:24



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: SPENCER, JOHN  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 04/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 04:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: DIAKOV

DATE/TIME OF TEST: 04/18/2018 05:07:59

---

Diagnostics Check	OK	05:01
Air Blank	0.000	05:01
Control Test	0.080	05:02
Air Blank	0.000	05:02
Subject Sample # 1	0.054	05:03
Breath Volume	1.515	05:03
Air Blank	0.000	05:04
Air Blank	0.000	05:06
Subject Sample # 2	0.053	05:06
Breath Volume	1.292	05:06
Air Blank	0.000	05:07
Control Test	0.080	05:07
Air Blank	0.000	05:07
Diagnostics Check	OK	05:07

---

Data Download Date/Time: 04/26/2018 14:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: RIVERA, VICTOR  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 04/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:06  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 04/19/2018 04:21:36

---

Diagnostics Check	OK	04:14
Air Blank	0.000	04:14
Control Test	0.080	04:14
Air Blank	0.000	04:15
Subject Sample # 1	0.060	04:17
Breath Volume	1.242	04:17
Air Blank	0.000	04:17
Air Blank	0.000	04:19
Subject Sample # 2	0.061	04:20
Breath Volume	1.300	04:20
Air Blank	0.000	04:20
Control Test	0.080	04:21
Air Blank	0.000	04:21
Diagnostics Check	OK	04:21

---

Data Download Date/Time: 04/26/2018 14:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: KELEMEN, JANOS  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 04/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: PEREZ DE JESUS

DATE/TIME OF TEST: 04/20/2018 01:53:20

---

Diagnostics Check	OK	01:45
Air Blank	0.000	01:45
Control Test	0.081	01:45
Air Blank	0.000	01:46
Subject Sample # 1	0.081	01:48
Breath Volume	1.222	01:48
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample # 2	0.086	01:51
Breath Volume	1.589	01:51
Air Blank	0.000	01:52
Control Test	0.080	01:52
Air Blank	0.000	01:53
Diagnostics Check	OK	01:53

---

Data Download Date/Time: 04/26/2018 14:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: GONCALVES,GUSTAVO  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 04/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: CARRIZO ROMERO

DATE/TIME OF TEST: 04/20/2018 03:44:44

---

Diagnostics Check	OK	03:36
Air Blank	0.000	03:36
Control Test	0.080	03:37
Air Blank	0.000	03:37
Subject Sample # 1	0.246	03:40
Breath Volume	1.511	03:40
Air Blank	0.000	03:40
Air Blank	0.000	03:42
Subject Sample # 2	0.251	03:43
Breath Volume	1.730	03:43
Air Blank	0.000	03:43
Control Test	0.078	03:44
Air Blank	0.000	03:44
Diagnostics Check	OK	03:44

---

Data Download Date/Time: 04/26/2018 14:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: MCWHERTER, PETER  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 04/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 06:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:59

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 04/25/2018 07:27:59

---

Diagnostics Check	OK	07:21
Air Blank	0.000	07:22
Control Test	0.081	07:22
Air Blank	0.000	07:22
Subject Sample # 1	0.102	07:23
Breath Volume	1.441	07:23
Air Blank	0.000	07:24
Air Blank	0.000	07:25
Subject Sample # 2	0.106	07:26
Breath Volume	1.234	07:26
Air Blank	0.000	07:27
Control Test	0.081	07:27
Air Blank	0.000	07:27
Diagnostics Check	OK	07:27

---

Data Download Date/Time: 04/26/2018 14:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: SPEERS, ANDREW  
ARRESTING OFFICER AGENCY: Maitland PD  
LAST AGENCY INSPECTION: 04/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: SHADDINGER

DATE/TIME OF TEST: 04/29/2018 01:03:52

---

Diagnostics Check	OK	00:57
Air Blank	0.000	00:57
Control Test	0.081	00:57
Air Blank	0.000	00:58
Subject Sample # 1	0.000	00:58
Breath Volume	1.566	00:58
Air Blank	0.000	00:59
Air Blank	0.000	01:01
Subject Sample # 2	0.000	01:02
Breath Volume	1.328	01:02
Air Blank	0.000	01:02
Control Test	0.081	01:03
Air Blank	0.000	01:03
Diagnostics Check	OK	01:03

---

Data Download Date/Time: 05/03/2018 16:55

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: FIGUEROA, HECTOR  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 04/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: CASTILLOHUETE

DATE/TIME OF TEST: 04/29/2018 03:44:14

---

Diagnostics Check		OK	03:36
Air Blank		0.000	03:36
Control Test		0.081	03:37
Air Blank		0.000	03:37
Subject Sample	# 1	0.208	03:39
Breath Volume		1.207	03:39
Air Blank		0.000	03:40
Air Blank		0.000	03:42
Subject Sample	# 2	0.219	03:42
Breath Volume		2.089	03:42
Air Blank		0.000	03:43
Control Test		0.080	03:43
Air Blank		0.000	03:44
Diagnostics Check		OK	03:44

---

Data Download Date/Time: 05/03/2018 16:55

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: ROBBINS,NATHAN  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 04/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: SANTIAGO RODRIGUEZ

DATE/TIME OF TEST: 04/29/2018 04:35:36

---

Diagnostics Check		OK	04:28
Air Blank		0.000	04:29
Control Test		0.081	04:29
Air Blank		0.000	04:30
Subject Sample	# 1	0.238	04:31
Breath Volume		2.289	04:31
Air Blank		0.000	04:31
Air Blank		0.000	04:33
Subject Sample	# 2	0.237	04:34
Breath Volume		3.085	04:34
Air Blank		0.000	04:34
Control Test		0.081	04:35
Air Blank		0.000	04:35
Diagnostics Check		OK	04:35

---

Data Download Date/Time: 05/03/2018 16:55



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: DAVILA, SAMUEL  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 04/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: NARCIS

DATE/TIME OF TEST: 04/30/2018 02:02:30

---

Diagnostics Check	OK	01:51		
Air Blank	0.000	01:51		
Control Test	0.081	01:51		
Air Blank	0.000	01:52		
Subject Sample	# 1	01:55	No Sample Provided	
Breath Volume	0.000	01:55		
Air Blank	0.000	01:55		
Air Blank	0.000	01:57		
Subject Sample	# 2	0.157	02:01	Volume Not Met
Breath Volume	0.585	02:01		
Air Blank	0.000	02:01		
Control Test	0.080	02:01		
Air Blank	0.000	02:02		
Diagnostics Check	OK	02:02		

---

Data Download Date/Time: 05/03/2018 16:55

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

INSTRUMENT REGISTERED TO: COCOA P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WANCA, WILLIAM, T  
BREATH TEST OPERATOR AGENCY: Rockledge PD  
ARRESTING OFFICER: HERNANDEZ, LUIS  
ARRESTING OFFICER AGENCY: Cocoa PD  
LAST AGENCY INSPECTION: 04/26/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A3  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:21  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: SPEARS

DATE/TIME OF TEST: 04/27/2018 03:07:20

---

Diagnostics Check	OK	03:00
Air Blank	0.000	03:00
Control Test	0.079	03:01
Air Blank	0.000	03:01
Subject Sample # 1	0.149	03:02
Breath Volume	1.296	03:02
Air Blank	0.000	03:03
Air Blank	0.000	03:05
Subject Sample # 2	0.145	03:05
Breath Volume	1.457	03:05
Air Blank	0.000	03:06
Control Test	0.078	03:06
Air Blank	0.000	03:07
Diagnostics Check	OK	03:07

---

Data Download Date/Time: 05/26/2018 20:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LYNCH,ROBERT,W  
BREATH TEST OPERATOR AGENCY: Rockledge PD  
ARRESTING OFFICER: HERBENER,TIMOTHY  
ARRESTING OFFICER AGENCY: Rockledge PD  
LAST AGENCY INSPECTION: 03/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 572920  
DRY GAS EXPIRY DATE: 09/13/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: CONKLIN

DATE/TIME OF TEST: 04/04/2018 03:09:36

---

Diagnostics Check	OK	03:06	
Air Blank	0.000	03:07	
Control Test	0.079	03:07	
Air Blank	0.000	03:07	
Subject Sample	# 1	03:08	Subject Test Refused
Breath Volume	0.000	03:08	
Air Blank	0.000	03:08	
Control Test	0.079	03:09	
Air Blank	0.000	03:09	
Diagnostics Check	OK	03:09	

---

Data Download Date/Time: 05/20/2018 09:35

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LYNCH,ROBERT,W  
BREATH TEST OPERATOR AGENCY: Rockledge PD  
ARRESTING OFFICER: LYNCH,ROBERT  
ARRESTING OFFICER AGENCY: Rockledge PD  
LAST AGENCY INSPECTION: 03/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 572920  
DRY GAS EXPIRY DATE: 09/13/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: THANG

DATE/TIME OF TEST: 04/07/2018 00:46:08

---

Diagnostics Check	OK	00:39
Air Blank	0.000	00:39
Control Test	0.079	00:40
Air Blank	0.000	00:40
Subject Sample # 1	0.229	00:41
Breath Volume	2.191	00:41
Air Blank	0.000	00:42
Air Blank	0.000	00:44
Subject Sample # 2	0.232	00:44
Breath Volume	1.742	00:44
Air Blank	0.000	00:45
Control Test	0.077	00:45
Air Blank	0.000	00:46
Diagnostics Check	OK	00:46

---

Data Download Date/Time: 05/20/2018 09:35

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LIVINGSTON,CHRISTOPHER,S  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: VAUGHN,JOHN  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 03/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814183  
DRY GAS EXPIRY DATE: 07/26/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: BERDAHL

DATE/TIME OF TEST: 04/11/2018 01:06:02

---

Diagnostics Check	OK	01:03	
Air Blank	0.000	01:04	
Control Test	0.079	01:04	
Air Blank		01:05	Purge Fail
Air Blank	0.000	01:06	

---

Data Download Date/Time: 04/30/2018 10:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LIVINGSTON,CHRISTOPHER,S  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: VAUGHN,JOHN  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 03/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814183  
DRY GAS EXPIRY DATE: 07/26/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: BERDAHL

DATE/TIME OF TEST: 04/11/2018 01:09:06

---

Diagnostics Check	OK	01:07	
Air Blank	0.000	01:07	
Control Test	0.069	01:08	Control Outside Tolerance
Air Blank	0.000	01:09	

---

Data Download Date/Time: 04/30/2018 10:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LIVINGSTON,CHRISTOPHER,S  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: VAUGHN,JOHN  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 03/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814183  
DRY GAS EXPIRY DATE: 07/26/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: BERDAHL

DATE/TIME OF TEST: 04/11/2018 01:16:47

---

Diagnostics Check		OK	01:10
Air Blank		0.000	01:11
Control Test		0.080	01:11
Air Blank		0.000	01:11
Subject Sample	# 1	0.000	01:12
Breath Volume		4.000	01:12
Air Blank		0.000	01:12
Air Blank		0.000	01:14
Subject Sample	# 2	0.000	01:15
Breath Volume		3.199	01:15
Air Blank		0.000	01:15
Control Test		0.078	01:16
Air Blank		0.000	01:16
Diagnostics Check		OK	01:16

---

Data Download Date/Time: 04/30/2018 10:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARLSON, EDWIN, J  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: WHITE, TAYLOR  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 03/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814183  
DRY GAS EXPIRY DATE: 07/26/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: FARMER

DATE/TIME OF TEST: 04/13/2018 02:32:35

---

Diagnostics Check	OK	02:26
Air Blank	0.000	02:26
Control Test	0.080	02:26
Air Blank	0.000	02:27
Subject Sample # 1	0.104	02:27
Breath Volume	3.046	02:27
Air Blank	0.000	02:28
Air Blank	0.000	02:30
Subject Sample # 2	0.098	02:30
Breath Volume	3.218	02:30
Air Blank	0.000	02:31
Control Test	0.080	02:31
Air Blank	0.000	02:32
Diagnostics Check	OK	02:32

---

Data Download Date/Time: 04/30/2018 10:39



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARLSON,EDWIN,J  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: CARLSON,EDWIN  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 03/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814183  
DRY GAS EXPIRY DATE: 07/26/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:01

SUBJECT'S LAST NAME: SIMS

DATE/TIME OF TEST: 04/16/2018 23:40:46

---

Diagnostics Check	OK	23:34
Air Blank	0.000	23:34
Control Test	0.080	23:35
Air Blank	0.000	23:35
Subject Sample # 1	0.227	23:36
Breath Volume	4.617	23:36
Air Blank	0.000	23:36
Air Blank	0.000	23:38
Subject Sample # 2	0.215	23:39
Breath Volume	4.660	23:39
Air Blank	0.000	23:39
Control Test	0.075	23:40
Air Blank	0.000	23:40
Diagnostics Check	OK	23:40

---

Data Download Date/Time: 04/30/2018 10:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARTHOLOMAY,RICHARD,J  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: BLACKMON,TERRANCE  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 03/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814183  
DRY GAS EXPIRY DATE: 07/26/2020  
VIOLATION CODE: DUI  
ARREST TIME: 16:49  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: VANEICK

DATE/TIME OF TEST: 04/22/2018 17:48:22

---

Diagnostics Check	OK	17:42
Air Blank	0.000	17:42
Control Test	0.079	17:43
Air Blank	0.000	17:43
Subject Sample # 1	0.000	17:44
Breath Volume	1.890	17:44
Air Blank	0.000	17:44
Air Blank	0.000	17:46
Subject Sample # 2	0.000	17:47
Breath Volume	1.691	17:47
Air Blank	0.000	17:47
Control Test	0.079	17:47
Air Blank	0.000	17:48
Diagnostics Check	OK	17:48

---

Data Download Date/Time: 04/30/2018 10:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FERREIRA,ZACHARY,J  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: FERREIRA,ZACHARY  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 03/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A5  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 05:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: STUTTE

DATE/TIME OF TEST: 04/01/2018 06:12:25

---

Diagnostics Check	OK	06:04
Air Blank	0.000	06:05
Control Test	0.080	06:05
Air Blank	0.000	06:06
Subject Sample # 1	0.170	06:06
Breath Volume	2.972	06:06
Air Blank	0.000	06:07
Air Blank	0.000	06:09
Subject Sample # 2	0.166	06:10
Breath Volume	1.863	06:10
Air Blank	0.000	06:11
Control Test	0.080	06:11
Air Blank	0.000	06:12
Diagnostics Check	OK	06:12

---

Data Download Date/Time: 04/26/2018 17:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WINSTEAD,JUSTIN,M  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: WINSTEAD,JUSTIN  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 03/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A5  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 03:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: STEINBACH

DATE/TIME OF TEST: 04/14/2018 04:37:58

---

Diagnostics Check	OK	04:30
Air Blank	0.000	04:31
Control Test	0.081	04:31
Air Blank	0.000	04:31
Subject Sample # 1	0.174	04:32
Breath Volume	1.761	04:32
Air Blank	0.000	04:33
Air Blank	0.000	04:34
Subject Sample # 2	0.164	04:36
Breath Volume	2.015	04:36
Air Blank	0.000	04:37
Control Test	0.079	04:37
Air Blank	0.000	04:37
Diagnostics Check	OK	04:37

---

Data Download Date/Time: 04/26/2018 17:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WINSTEAD,JUSTIN,M  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: WINSTEAD,JUSTIN  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 03/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A5  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 05:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:05

SUBJECT'S LAST NAME: HAMBLIN

DATE/TIME OF TEST: 04/20/2018 06:44:09

---

Diagnostics Check	OK	06:36
Air Blank	0.000	06:37
Control Test	0.081	06:37
Air Blank	0.000	06:38
Subject Sample # 1	0.000	06:39
Breath Volume	1.144	06:39
Air Blank	0.000	06:40
Air Blank	0.000	06:42
Subject Sample # 2	0.000	06:42
Breath Volume	2.507	06:42
Air Blank	0.000	06:43
Control Test	0.081	06:43
Air Blank	0.000	06:44
Diagnostics Check	OK	06:44

---

Data Download Date/Time: 04/26/2018 17:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BIVINS, HEATHER, C  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: WEIDNER, CODY  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 03/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A5  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 22:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: ERNST

DATE/TIME OF TEST: 04/23/2018 23:25:10

---

Diagnostics Check	OK	23:18
Air Blank	0.000	23:19
Control Test	0.080	23:19
Air Blank	0.000	23:20
Subject Sample # 1	0.159	23:20
Breath Volume	1.937	23:20
Air Blank	0.000	23:21
Air Blank	0.000	23:23
Subject Sample # 2	0.161	23:23
Breath Volume	1.914	23:23
Air Blank	0.000	23:24
Control Test	0.080	23:24
Air Blank	0.000	23:25
Diagnostics Check	OK	23:25

---

Data Download Date/Time: 04/26/2018 17:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ,DAILYN,  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: NUTTING,KAYLEE  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 03/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A5  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 04/25/2018 02:24:10

---

Diagnostics Check	OK	02:17
Air Blank	0.000	02:18
Control Test	0.081	02:18
Air Blank	0.000	02:18
Subject Sample # 1	0.226	02:19
Breath Volume	1.933	02:19
Air Blank	0.000	02:20
Air Blank	0.000	02:21
Subject Sample # 2	0.214	02:22
Breath Volume	2.300	02:22
Air Blank	0.000	02:23
Control Test	0.081	02:23
Air Blank	0.000	02:24
Diagnostics Check	OK	02:24

---

Data Download Date/Time: 04/26/2018 17:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROWELL,NATHANAEL,A  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: ROWELL,NATHANAEL  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 04/26/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A5  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 22:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: BEAULIEU

DATE/TIME OF TEST: 04/27/2018 23:34:30

---

Diagnostics Check	OK	23:28
Air Blank	0.000	23:28
Control Test	0.080	23:28
Air Blank	0.000	23:29
Subject Sample # 1	0.137	23:29
Breath Volume	4.175	23:29
Air Blank	0.000	23:30
Air Blank	0.000	23:32
Subject Sample # 2	0.137	23:32
Breath Volume	4.593	23:32
Air Blank	0.000	23:33
Control Test	0.079	23:33
Air Blank	0.000	23:34
Diagnostics Check	OK	23:34

---

Data Download Date/Time: 05/23/2018 18:25



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BIVINS,HEATHER,C  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: BIVINS,HEATHER  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 04/26/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A5  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 23:49  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 04/30/2018 02:57:51

---

Diagnostics Check	OK	02:49
Air Blank	0.000	02:49
Control Test	0.081	02:50
Air Blank	0.000	02:50
Subject Sample # 1	0.000	02:52
Breath Volume	1.242	02:52
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample # 2	0.000	02:56
Breath Volume	1.402	02:56
Air Blank	0.000	02:56
Control Test	0.081	02:57
Air Blank	0.000	02:57
Diagnostics Check	OK	02:57

---

Data Download Date/Time: 05/23/2018 18:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROLPH,AMANDA,L  
BREATH TEST OPERATOR AGENCY: Palm Bay PD  
ARRESTING OFFICER: ROLPH,AMANDA  
ARRESTING OFFICER AGENCY: Palm Bay PD  
LAST AGENCY INSPECTION: 03/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:44  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: VANCE

DATE/TIME OF TEST: 04/07/2018 04:25:24

---

Diagnostics Check	OK	04:18
Air Blank	0.000	04:19
Control Test	0.082	04:19
Air Blank	0.000	04:20
Subject Sample # 1	0.113	04:20
Breath Volume	2.828	04:20
Air Blank	0.000	04:21
Air Blank	0.000	04:23
Subject Sample # 2	0.113	04:23
Breath Volume	2.527	04:23
Air Blank	0.000	04:24
Control Test	0.080	04:24
Air Blank	0.000	04:25
Diagnostics Check	OK	04:25

---

Data Download Date/Time: 04/25/2018 20:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 03/28/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 22817080A4

DRY GAS EXPIRY DATE: 10/05/2019

VIOLATION CODE: DUI

ARREST TIME: 19:25

GENDER: Female

OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: MASI

DATE/TIME OF TEST: 04/09/2018 19:34:46

---

Diagnostics Check	OK	19:28
Air Blank	0.000	19:29
Control Test	0.080	19:29
Air Blank	0.000	19:29
Subject Sample # 1	0.178	19:30
Breath Volume	2.000	19:30
Air Blank	0.000	19:30
Air Blank	0.000	19:32
Subject Sample # 2	0.172	19:33
Breath Volume	1.851	19:33
Air Blank	0.000	19:33
Control Test	0.080	19:34
Air Blank	0.000	19:34
Diagnostics Check	OK	19:34

---

Data Download Date/Time: 04/25/2018 20:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LUTZ,EDWIN,J  
BREATH TEST OPERATOR AGENCY: Palm Bay PD  
ARRESTING OFFICER: LUTZ,EDWIN  
ARRESTING OFFICER AGENCY: Palm Bay PD  
LAST AGENCY INSPECTION: 03/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 04/10/2018 01:53:18

---

Diagnostics Check	OK	01:46
Air Blank	0.000	01:46
Control Test	0.080	01:47
Air Blank	0.000	01:47
Subject Sample # 1	0.053	01:48
Breath Volume	1.652	01:48
Air Blank	0.000	01:49
Air Blank	0.000	01:50
Subject Sample # 2	0.058	01:51
Breath Volume	1.761	01:51
Air Blank	0.000	01:52
Control Test	0.079	01:52
Air Blank	0.000	01:53
Diagnostics Check	OK	01:53

---

Data Download Date/Time: 04/25/2018 20:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 03/28/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 22817080A4

DRY GAS EXPIRY DATE: 10/05/2019

VIOLATION CODE: DUI

ARREST TIME: 22:43

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 04/10/2018 22:52:48

---

Diagnostics Check	OK	22:46
Air Blank	0.000	22:47
Control Test	0.081	22:47
Air Blank	0.000	22:47
Subject Sample # 1	0.167	22:48
Breath Volume	2.406	22:48
Air Blank	0.000	22:49
Air Blank	0.000	22:50
Subject Sample # 2	0.158	22:51
Breath Volume	1.871	22:51
Air Blank	0.000	22:51
Control Test	0.079	22:52
Air Blank	0.000	22:52
Diagnostics Check	OK	22:52

---

Data Download Date/Time: 04/25/2018 20:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LUTZ,EDWIN,J  
BREATH TEST OPERATOR AGENCY: Palm Bay PD  
ARRESTING OFFICER: LUTZ,EDWIN  
ARRESTING OFFICER AGENCY: Palm Bay PD  
LAST AGENCY INSPECTION: 03/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:33  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: SWIFT

DATE/TIME OF TEST: 04/15/2018 01:54:19

---

Diagnostics Check	OK	01:48
Air Blank	0.000	01:48
Control Test	0.079	01:49
Air Blank	0.000	01:49
Subject Sample # 1	0.072	01:49
Breath Volume	2.308	01:49
Air Blank	0.000	01:50
Air Blank	0.000	01:52
Subject Sample # 2	0.073	01:52
Breath Volume	1.871	01:52
Air Blank	0.000	01:53
Control Test	0.081	01:53
Air Blank	0.000	01:54
Diagnostics Check	OK	01:54

---

Data Download Date/Time: 04/25/2018 20:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LUTZ,EDWIN,J  
BREATH TEST OPERATOR AGENCY: Palm Bay PD  
ARRESTING OFFICER: LUTZ,EDWIN  
ARRESTING OFFICER AGENCY: Palm Bay PD  
LAST AGENCY INSPECTION: 03/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 04:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:03

SUBJECT'S LAST NAME: RAY

DATE/TIME OF TEST: 04/15/2018 05:53:36

---

Diagnostics Check		OK	05:46
Air Blank		0.000	05:47
Control Test		0.080	05:47
Air Blank		0.000	05:47
Subject Sample	# 1	0.187	05:48
Breath Volume		1.785	05:48
Air Blank		0.000	05:49
Air Blank		0.000	05:51
Subject Sample	# 2	0.190	05:52
Breath Volume		1.328	05:52
Air Blank		0.000	05:52
Control Test		0.080	05:53
Air Blank		0.000	05:53
Diagnostics Check		OK	05:53

---

Data Download Date/Time: 04/25/2018 20:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LUTZ,EDWIN,J  
BREATH TEST OPERATOR AGENCY: Palm Bay PD  
ARRESTING OFFICER: LUTZ,EDWIN  
ARRESTING OFFICER AGENCY: Palm Bay PD  
LAST AGENCY INSPECTION: 03/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: MCGRAW

DATE/TIME OF TEST: 04/22/2018 01:02:36

---

Diagnostics Check		OK	00:56
Air Blank		0.000	00:56
Control Test		0.083	00:57
Air Blank		0.000	00:57
Subject Sample	# 1	0.048	00:58
Breath Volume		1.410	00:58
Air Blank		0.000	00:58
Air Blank		0.000	01:00
Subject Sample	# 2	0.052	01:01
Breath Volume		1.945	01:01
Air Blank		0.000	01:01
Control Test		0.080	01:02
Air Blank		0.000	01:02
Diagnostics Check		OK	01:02

---

Data Download Date/Time: 04/25/2018 20:33



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LUTZ,EDWIN,J  
BREATH TEST OPERATOR AGENCY: Palm Bay PD  
ARRESTING OFFICER: LUTZ,EDWIN  
ARRESTING OFFICER AGENCY: Palm Bay PD  
LAST AGENCY INSPECTION: 03/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: FLOHR

DATE/TIME OF TEST: 04/22/2018 02:09:22

---

Diagnostics Check	OK	02:02
Air Blank	0.000	02:03
Control Test	0.081	02:03
Air Blank	0.000	02:04
Subject Sample # 1	0.159	02:04
Breath Volume	2.250	02:04
Air Blank	0.000	02:05
Air Blank	0.000	02:07
Subject Sample # 2	0.159	02:07
Breath Volume	1.695	02:07
Air Blank	0.000	02:08
Control Test	0.080	02:08
Air Blank	0.000	02:09
Diagnostics Check	OK	02:09

---

Data Download Date/Time: 04/25/2018 20:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LUTZ,EDWIN,J  
BREATH TEST OPERATOR AGENCY: Palm Bay PD  
ARRESTING OFFICER: LUTZ,EDWIN  
ARRESTING OFFICER AGENCY: Palm Bay PD  
LAST AGENCY INSPECTION: 03/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: DENNY JR

DATE/TIME OF TEST: 04/22/2018 22:22:20

---

Diagnostics Check	OK	22:16
Air Blank	0.000	22:16
Control Test	0.080	22:16
Air Blank	0.000	22:17
Subject Sample # 1	0.141	22:17
Breath Volume	2.507	22:17
Air Blank	0.000	22:18
Air Blank	0.000	22:20
Subject Sample # 2	0.139	22:20
Breath Volume	1.949	22:20
Air Blank	0.000	22:21
Control Test	0.079	22:21
Air Blank	0.000	22:22
Diagnostics Check	OK	22:22

---

Data Download Date/Time: 04/25/2018 20:33

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 04/25/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 22817080A4

DRY GAS EXPIRY DATE: 10/05/2019

VIOLATION CODE: DUI

ARREST TIME: 20:48

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:32

SUBJECT'S LAST NAME: SUTTON

DATE/TIME OF TEST: 04/26/2018 20:58:43

---

Diagnostics Check		OK	20:52
Air Blank		0.000	20:53
Control Test		0.081	20:53
Air Blank		0.000	20:53
Subject Sample	# 1	0.196	20:54
Breath Volume		2.691	20:54
Air Blank		0.000	20:54
Air Blank		0.000	20:56
Subject Sample	# 2	0.200	20:57
Breath Volume		2.582	20:57
Air Blank		0.000	20:57
Control Test		0.080	20:58
Air Blank		0.000	20:58
Diagnostics Check		OK	20:58

---

Data Download Date/Time: 05/25/2018 18:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 04/25/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 22817080A4

DRY GAS EXPIRY DATE: 10/05/2019

VIOLATION CODE: DUI

ARREST TIME: 23:05

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 04/27/2018 23:15:38

---

Diagnostics Check	OK	23:09
Air Blank	0.000	23:09
Control Test	0.082	23:10
Air Blank	0.000	23:10
Subject Sample # 1	0.119	23:11
Breath Volume	2.265	23:11
Air Blank	0.000	23:11
Air Blank	0.000	23:13
Subject Sample # 2	0.120	23:14
Breath Volume	2.101	23:14
Air Blank	0.000	23:14
Control Test	0.078	23:15
Air Blank	0.000	23:15
Diagnostics Check	OK	23:15

---

Data Download Date/Time: 05/25/2018 18:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 04/25/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 22817080A4

DRY GAS EXPIRY DATE: 10/05/2019

VIOLATION CODE: DUI

ARREST TIME: 01:58

GENDER: Female

OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: MCCLELLAN

DATE/TIME OF TEST: 04/28/2018 02:16:30

---

Diagnostics Check		OK	02:10
Air Blank		0.000	02:11
Control Test		0.083	02:11
Air Blank		0.000	02:11
Subject Sample	# 1	0.164	02:12
Breath Volume		3.152	02:12
Air Blank		0.000	02:12
Air Blank		0.000	02:14
Subject Sample	# 2	0.161	02:14
Breath Volume		3.109	02:14
Air Blank		0.000	02:15
Control Test		0.081	02:15
Air Blank		0.000	02:16
Diagnostics Check		OK	02:16

---

Data Download Date/Time: 05/25/2018 18:38

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LUTZ,EDWIN,J  
BREATH TEST OPERATOR AGENCY: Palm Bay PD  
ARRESTING OFFICER: LUTZ,EDWIN  
ARRESTING OFFICER AGENCY: Palm Bay PD  
LAST AGENCY INSPECTION: 04/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: ZACCAGNINI

DATE/TIME OF TEST: 04/29/2018 22:15:05

---

Diagnostics Check		OK	22:08
Air Blank		0.000	22:09
Control Test		0.081	22:09
Air Blank		0.000	22:09
Subject Sample	# 1	0.162	22:10
Breath Volume		2.839	22:10
Air Blank		0.000	22:10
Air Blank		0.000	22:12
Subject Sample	# 2	0.164	22:13
Breath Volume		2.992	22:13
Air Blank		0.000	22:14
Control Test		0.081	22:14
Air Blank		0.000	22:15
Diagnostics Check		OK	22:15

---

Data Download Date/Time: 05/25/2018 18:38

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAMILTON,J,W  
BREATH TEST OPERATOR AGENCY: Palm Bay PD  
ARRESTING OFFICER: HAMILTON,J  
ARRESTING OFFICER AGENCY: Palm Bay PD  
LAST AGENCY INSPECTION: 03/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: PAGAN TORRES

DATE/TIME OF TEST: 04/10/2018 01:02:58

---

Diagnostics Check		OK	00:56
Air Blank		0.000	00:56
Control Test		0.078	00:56
Air Blank		0.000	00:57
Subject Sample	# 1	0.172	00:57
Breath Volume		3.027	00:57
Air Blank		0.000	00:58
Air Blank		0.000	01:00
Subject Sample	# 2	0.164	01:01
Breath Volume		3.472	01:01
Air Blank		0.000	01:02
Control Test		0.078	01:02
Air Blank		0.000	01:02
Diagnostics Check		OK	01:02

---

Data Download Date/Time: 04/25/2018 20:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAMILTON,J,W  
BREATH TEST OPERATOR AGENCY: Palm Bay PD  
ARRESTING OFFICER: HAMILTON,J  
ARRESTING OFFICER AGENCY: Palm Bay PD  
LAST AGENCY INSPECTION: 03/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: COLON SANCHEZ

DATE/TIME OF TEST: 04/11/2018 03:18:33

---

Diagnostics Check	OK	03:09	
Air Blank	0.000	03:09	
Control Test	0.080	03:10	
Air Blank	0.000	03:10	
Subject Sample	# 1	03:13	No Sample Provided
Breath Volume	0.000	03:13	
Air Blank	0.000	03:14	
Air Blank	0.000	03:16	
Subject Sample	# 2	03:16	Subject Test Refused
Breath Volume	0.078	03:16	
Air Blank	0.000	03:17	
Control Test	0.080	03:17	
Air Blank	0.000	03:18	
Diagnostics Check	Fail	03:18	Voltage/Current Test
Air Blank		03:18	RFI Detect

---

Data Download Date/Time: 04/25/2018 20:23



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAMILTON,J,W  
BREATH TEST OPERATOR AGENCY: Palm Bay PD  
ARRESTING OFFICER: VALENTIN,C  
ARRESTING OFFICER AGENCY: Palm Bay PD  
LAST AGENCY INSPECTION: 03/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 04/15/2018 02:20:17

---

Diagnostics Check	OK	02:10	
Air Blank	0.000	02:11	
Control Test	0.080	02:11	
Air Blank	0.000	02:11	
Subject Sample # 1	0.246	02:13	
Breath Volume	1.132	02:13	
Air Blank	0.000	02:13	
Air Blank	0.000	02:15	
Subject Sample # 2	0.259	02:18	
Breath Volume	2.867	02:18	
Air Blank	0.000	02:18	
Control Test	0.079	02:19	
Air Blank	0.000	02:19	
Diagnostics Check	Fail	02:19	Analytical Stability Test
Air Blank	0.000	02:20	

---

Data Download Date/Time: 04/25/2018 20:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAMILTON,J,W  
BREATH TEST OPERATOR AGENCY: Palm Bay PD  
ARRESTING OFFICER: HAMILTON,J  
ARRESTING OFFICER AGENCY: Palm Bay PD  
LAST AGENCY INSPECTION: 03/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: BOGUS

DATE/TIME OF TEST: 04/16/2018 02:20:13

---

Diagnostics Check	OK	02:12
Air Blank	0.000	02:13
Control Test	0.081	02:13
Air Blank	0.000	02:14
Subject Sample # 1	0.173	02:15
Breath Volume	1.843	02:15
Air Blank	0.000	02:16
Air Blank	0.000	02:18
Subject Sample # 2	0.172	02:18
Breath Volume	2.417	02:18
Air Blank	0.000	02:19
Control Test	0.080	02:19
Air Blank	0.000	02:20
Diagnostics Check	OK	02:20

---

Data Download Date/Time: 04/25/2018 20:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAMILTON,J,W  
BREATH TEST OPERATOR AGENCY: Palm Bay PD  
ARRESTING OFFICER: HAMILTON,J  
ARRESTING OFFICER AGENCY: Palm Bay PD  
LAST AGENCY INSPECTION: 03/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 18:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:52

SUBJECT'S LAST NAME: GUNTHER

DATE/TIME OF TEST: 04/21/2018 19:26:48

---

Diagnostics Check	OK	19:20
Air Blank	0.000	19:21
Control Test	0.079	19:21
Air Blank	0.000	19:22
Subject Sample # 1	0.205	19:22
Breath Volume	2.128	19:22
Air Blank	0.000	19:23
Air Blank	0.000	19:24
Subject Sample # 2	0.200	19:25
Breath Volume	2.457	19:25
Air Blank	0.000	19:25
Control Test	0.079	19:26
Air Blank	0.000	19:26
Diagnostics Check	OK	19:26

---

Data Download Date/Time: 04/25/2018 20:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HAMILTON,J,W

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: HAMILTON,J

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 04/25/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 22817080A4

DRY GAS EXPIRY DATE: 10/05/2019

VIOLATION CODE: DUI

ARREST TIME: 02:40

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: DOLBOW

DATE/TIME OF TEST: 04/29/2018 03:02:43

---

Diagnostics Check	OK	02:54
Air Blank	0.000	02:55
Control Test	0.080	02:55
Air Blank	0.000	02:55
Subject Sample # 1	0.149	02:58
Breath Volume	2.582	02:58
Air Blank	0.000	02:58
Air Blank	0.000	03:00
Subject Sample # 2	0.138	03:01
Breath Volume	1.515	03:01
Air Blank	0.000	03:01
Control Test	0.080	03:02
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

---

Data Download Date/Time: 05/25/2018 18:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HAMILTON,J,W

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: HAMILTON,J

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 04/25/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 22817080A4

DRY GAS EXPIRY DATE: 10/05/2019

VIOLATION CODE: DUI

ARREST TIME: 20:15

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:08

SUBJECT'S LAST NAME: LENT

DATE/TIME OF TEST: 04/29/2018 20:36:32

---

Diagnostics Check	OK	20:28
Air Blank	0.000	20:29
Control Test	0.081	20:29
Air Blank	0.000	20:29
Subject Sample # 1	0.191	20:31
Breath Volume	3.335	20:31
Air Blank	0.000	20:31
Air Blank	0.000	20:33
Subject Sample # 2	0.190	20:35
Breath Volume	3.433	20:35
Air Blank	0.000	20:35
Control Test	0.078	20:35
Air Blank	0.000	20:36
Diagnostics Check	OK	20:36

---

Data Download Date/Time: 05/25/2018 18:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BETHAM,KEITH,D  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: MONANI,MICHAEL  
ARRESTING OFFICER AGENCY: Oviedo PD  
LAST AGENCY INSPECTION: 03/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: MOORHEAD

DATE/TIME OF TEST: 04/01/2018 04:29:55

---

Diagnostics Check	OK	04:23
Air Blank	0.000	04:24
Control Test	0.080	04:24
Air Blank	0.000	04:24
Subject Sample # 1	0.124	04:25
Breath Volume	3.039	04:25
Air Blank	0.000	04:26
Air Blank	0.000	04:27
Subject Sample # 2	0.121	04:28
Breath Volume	2.925	04:28
Air Blank	0.000	04:29
Control Test	0.079	04:29
Air Blank	0.000	04:29
Diagnostics Check	OK	04:29

---

Data Download Date/Time: 04/29/2018 01:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARCIA, RAY,  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: ROGAN, BRITTANY  
ARRESTING OFFICER AGENCY: Winter Springs PD  
LAST AGENCY INSPECTION: 03/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:24  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: BOURNE

DATE/TIME OF TEST: 04/01/2018 22:18:52

---

Diagnostics Check	OK	22:12
Air Blank	0.000	22:12
Control Test	0.079	22:12
Air Blank	0.000	22:13
Subject Sample # 1	0.172	22:13
Breath Volume	2.093	22:13
Air Blank	0.000	22:14
Air Blank	0.000	22:16
Subject Sample # 2	0.175	22:17
Breath Volume	2.199	22:17
Air Blank	0.000	22:17
Control Test	0.079	22:18
Air Blank	0.000	22:18
Diagnostics Check	OK	22:18

---

Data Download Date/Time: 04/29/2018 01:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARCIA, RAY,  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: MOREL, ANTHONY  
ARRESTING OFFICER AGENCY: Winter Springs PD  
LAST AGENCY INSPECTION: 03/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: VAUGHN

DATE/TIME OF TEST: 04/02/2018 01:32:04

---

Diagnostics Check	OK	01:25
Air Blank	0.000	01:26
Control Test	0.080	01:26
Air Blank	0.000	01:27
Subject Sample # 1	0.209	01:27
Breath Volume	3.023	01:27
Air Blank	0.000	01:28
Air Blank	0.000	01:30
Subject Sample # 2	0.213	01:30
Breath Volume	2.734	01:30
Air Blank	0.000	01:31
Control Test	0.079	01:31
Air Blank	0.000	01:31
Diagnostics Check	OK	01:32

---

Data Download Date/Time: 04/29/2018 01:47



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARCIA, RAY,  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: PRICE, ZACHARY  
ARRESTING OFFICER AGENCY: Casselberry PD  
LAST AGENCY INSPECTION: 03/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: AN

DATE/TIME OF TEST: 04/05/2018 04:19:04

---

Diagnostics Check	OK	04:13
Air Blank	0.000	04:13
Control Test	0.081	04:13
Air Blank	0.000	04:14
Subject Sample # 1	0.179	04:14
Breath Volume	1.925	04:14
Air Blank	0.000	04:15
Air Blank	0.000	04:17
Subject Sample # 2	0.176	04:17
Breath Volume	1.656	04:17
Air Blank	0.000	04:18
Control Test	0.080	04:18
Air Blank	0.000	04:18
Diagnostics Check	OK	04:19

---

Data Download Date/Time: 04/29/2018 01:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BETHAM,KEITH,D  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: BOUTELLE,ADAM  
ARRESTING OFFICER AGENCY: Sanford PD  
LAST AGENCY INSPECTION: 03/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:42  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: SHAVER

DATE/TIME OF TEST: 04/06/2018 02:38:51

---

Diagnostics Check		OK	02:27
Air Blank		0.000	02:27
Control Test		0.081	02:27
Air Blank		0.000	02:28
Subject Sample	# 1	0.229	02:29
Breath Volume		1.765	02:29
Air Blank		0.000	02:29
Air Blank		0.000	02:31
Subject Sample	# 2	0.206	02:34
Breath Volume		1.144	02:34
Air Blank		0.000	02:34
Air Blank		0.000	02:36
Subject Sample	# 3	0.209	02:37
Breath Volume		1.398	02:37
Air Blank		0.000	02:37
Control Test		0.078	02:38
Air Blank		0.000	02:38
Diagnostics Check		OK	02:38

---

Data Download Date/Time: 04/29/2018 01:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BETHAM,KEITH,D  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: ANCION,CALIX  
ARRESTING OFFICER AGENCY: Seminole County SO  
LAST AGENCY INSPECTION: 03/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: POPE

DATE/TIME OF TEST: 04/06/2018 03:36:50

---

Diagnostics Check	OK	03:30
Air Blank	0.000	03:31
Control Test	0.081	03:31
Air Blank	0.000	03:32
Subject Sample # 1	0.000	03:32
Breath Volume	2.121	03:32
Air Blank	0.000	03:32
Air Blank	0.000	03:34
Subject Sample # 2	0.000	03:35
Breath Volume	1.238	03:35
Air Blank	0.000	03:35
Control Test	0.080	03:36
Air Blank	0.000	03:36
Diagnostics Check	OK	03:36

---

Data Download Date/Time: 04/29/2018 01:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEADOWS,SHAWN,M  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: QUINN,GREGORY  
ARRESTING OFFICER AGENCY: Sanford PD  
LAST AGENCY INSPECTION: 03/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 19:28  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:16

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 04/06/2018 21:51:56

---

Diagnostics Check	OK	21:42
Air Blank	0.000	21:42
Control Test	0.080	21:43
Air Blank	0.000	21:43
Subject Sample # 1	0.211	21:45
Breath Volume	1.761	21:45
Air Blank	0.000	21:46
Air Blank	0.000	21:48
Subject Sample # 2	0.219	21:50
Breath Volume	1.160	21:50
Air Blank	0.000	21:50
Control Test	0.078	21:51
Air Blank	0.000	21:51
Diagnostics Check	OK	21:51

---

Data Download Date/Time: 04/29/2018 01:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BETHAM,KEITH,D  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: KINKOPF,TIM  
ARRESTING OFFICER AGENCY: Seminole County SO  
LAST AGENCY INSPECTION: 03/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: TREMBLAY

DATE/TIME OF TEST: 04/07/2018 02:19:11

---

Diagnostics Check	OK	02:12
Air Blank	0.000	02:13
Control Test	0.080	02:13
Air Blank	0.000	02:13
Subject Sample # 1	0.111	02:14
Breath Volume	3.289	02:14
Air Blank	0.000	02:15
Air Blank	0.000	02:17
Subject Sample # 2	0.109	02:17
Breath Volume	3.148	02:17
Air Blank	0.000	02:18
Control Test	0.079	02:18
Air Blank	0.000	02:19
Diagnostics Check	OK	02:19

---

Data Download Date/Time: 04/29/2018 01:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BETHAM,KEITH,D  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: JONES,ANDREW  
ARRESTING OFFICER AGENCY: Winter Springs PD  
LAST AGENCY INSPECTION: 03/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: BALMAT

DATE/TIME OF TEST: 04/07/2018 02:45:13

---

Diagnostics Check	OK	02:38
Air Blank	0.000	02:39
Control Test	0.080	02:39
Air Blank	0.000	02:40
Subject Sample # 1	0.080	02:40
Breath Volume	2.867	02:40
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.083	02:43
Breath Volume	2.945	02:43
Air Blank	0.000	02:44
Control Test	0.080	02:44
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

---

Data Download Date/Time: 04/29/2018 01:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: MEADOWS,SHAWN  
ARRESTING OFFICER AGENCY: Seminole County SO  
LAST AGENCY INSPECTION: 03/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: JEREZ

DATE/TIME OF TEST: 04/07/2018 03:19:16

---

Diagnostics Check	OK	03:12
Air Blank	0.000	03:13
Control Test	0.079	03:13
Air Blank	0.000	03:13
Subject Sample # 1	0.208	03:14
Breath Volume	3.277	03:14
Air Blank	0.000	03:15
Air Blank	0.000	03:17
Subject Sample # 2	0.223	03:17
Breath Volume	3.707	03:17
Air Blank	0.000	03:18
Control Test	0.078	03:18
Air Blank	0.000	03:19
Diagnostics Check	OK	03:19

---

Data Download Date/Time: 04/29/2018 01:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: PIERCE,BRIAN  
ARRESTING OFFICER AGENCY: Sanford PD  
LAST AGENCY INSPECTION: 03/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: DAVIS JR

DATE/TIME OF TEST: 04/07/2018 22:39:19

---

Diagnostics Check	OK	22:32
Air Blank	0.000	22:33
Control Test	0.081	22:33
Air Blank	0.000	22:34
Subject Sample # 1	0.184	22:34
Breath Volume	2.414	22:34
Air Blank	0.000	22:35
Air Blank	0.000	22:37
Subject Sample # 2	0.177	22:37
Breath Volume	2.378	22:37
Air Blank	0.000	22:38
Control Test	0.079	22:38
Air Blank	0.000	22:39
Diagnostics Check	OK	22:39

---

Data Download Date/Time: 04/29/2018 01:47



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: KINKOPF,TIM  
ARRESTING OFFICER AGENCY: Seminole County SO  
LAST AGENCY INSPECTION: 03/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: TOMMEY

DATE/TIME OF TEST: 04/08/2018 04:56:06

---

Diagnostics Check	OK	04:48	
Air Blank	0.000	04:49	
Control Test	0.080	04:49	
Air Blank	0.000	04:50	
Subject Sample # 1	0.015	04:51	
Breath Volume	3.289	04:51	
Air Blank	0.000	04:51	
Air Blank	0.000	04:53	
Subject Sample # 2	0.010	04:54	
Breath Volume	1.785	04:54	
Air Blank	0.000	04:54	
Control Test	0.078	04:54	
Air Blank		04:55	Ambient Fail
Air Blank	0.000	04:56	

---

Data Download Date/Time: 04/29/2018 01:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: KINKOPF,TIM  
ARRESTING OFFICER AGENCY: Seminole County SO  
LAST AGENCY INSPECTION: 03/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: TOMMEY

DATE/TIME OF TEST: 04/08/2018 05:04:11

---

Diagnostics Check	OK	04:58	
Air Blank	0.000	04:58	
Control Test	0.077	04:59	
Air Blank	0.000	04:59	
Subject Sample # 1	0.009	05:00	
Breath Volume	2.156	05:00	
Air Blank	0.000	05:00	
Air Blank	0.000	05:02	
Subject Sample # 2	0.000	05:02	
Breath Volume	1.960	05:02	
Air Blank	0.000	05:03	
Control Test	0.074	05:03	Control Outside Tolerance
Air Blank	0.000	05:04	

---

Data Download Date/Time: 04/29/2018 01:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: HUNDEMANN,ROBERT  
ARRESTING OFFICER AGENCY: Sanford PD  
LAST AGENCY INSPECTION: 03/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 04:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: LAMM

DATE/TIME OF TEST: 04/08/2018 05:34:43

---

Diagnostics Check	OK	05:28
Air Blank	0.000	05:28
Control Test	0.079	05:29
Air Blank	0.000	05:29
Subject Sample # 1	0.155	05:30
Breath Volume	3.449	05:30
Air Blank	0.000	05:30
Air Blank	0.000	05:32
Subject Sample # 2	0.152	05:33
Breath Volume	3.316	05:33
Air Blank	0.000	05:33
Control Test	0.078	05:34
Air Blank	0.000	05:34
Diagnostics Check	OK	05:34

---

Data Download Date/Time: 04/29/2018 01:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARCIA, RAY,  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: WOODARD, DALLESSANDRO  
ARRESTING OFFICER AGENCY: Winter Springs PD  
LAST AGENCY INSPECTION: 03/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 04:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:02

SUBJECT'S LAST NAME: MCLEROY

DATE/TIME OF TEST: 04/12/2018 05:41:30

---

Diagnostics Check		OK	05:34
Air Blank		0.000	05:35
Control Test		0.080	05:35
Air Blank		0.000	05:35
Subject Sample	# 1	0.267	05:36
Breath Volume		2.011	05:36
Air Blank		0.000	05:37
Air Blank		0.000	05:39
Subject Sample	# 2	0.266	05:39
Breath Volume		1.812	05:39
Air Blank		0.000	05:40
Control Test		0.079	05:40
Air Blank		0.000	05:41
Diagnostics Check		OK	05:41

---

Data Download Date/Time: 04/29/2018 01:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BETHAM,KEITH,D  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: COPPOLA,TYSON  
ARRESTING OFFICER AGENCY: Longwood PD  
LAST AGENCY INSPECTION: 03/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: RAMOS-LOPEZ

DATE/TIME OF TEST: 04/14/2018 22:23:05

---

Diagnostics Check		OK	22:16
Air Blank		0.000	22:16
Control Test		0.080	22:17
Air Blank		0.000	22:17
Subject Sample	# 1	0.158	22:18
Breath Volume		2.656	22:18
Air Blank		0.000	22:18
Air Blank		0.000	22:20
Subject Sample	# 2	0.161	22:21
Breath Volume		2.335	22:21
Air Blank		0.000	22:22
Control Test		0.081	22:22
Air Blank		0.000	22:23
Diagnostics Check		OK	22:23

---

Data Download Date/Time: 04/29/2018 01:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BETHAM,KEITH,D  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: MOREL,ANTHONY  
ARRESTING OFFICER AGENCY: Winter Springs PD  
LAST AGENCY INSPECTION: 03/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: AUSTIN

DATE/TIME OF TEST: 04/15/2018 01:17:45

---

Diagnostics Check		OK	01:11
Air Blank		0.000	01:11
Control Test		0.080	01:11
Air Blank		0.000	01:12
Subject Sample	# 1	0.245	01:12
Breath Volume		2.871	01:12
Air Blank		0.000	01:13
Air Blank		0.000	01:15
Subject Sample	# 2	0.236	01:16
Breath Volume		2.824	01:16
Air Blank		0.000	01:16
Control Test		0.079	01:17
Air Blank		0.000	01:17
Diagnostics Check		OK	01:17

---

Data Download Date/Time: 04/29/2018 01:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COLBURN,MEGAN,E  
BREATH TEST OPERATOR AGENCY: Casselberry PD  
ARRESTING OFFICER: COLBURN,MEGAN  
ARRESTING OFFICER AGENCY: Casselberry PD  
LAST AGENCY INSPECTION: 03/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 17:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 04/15/2018 18:46:08

---

Diagnostics Check	OK	18:39
Air Blank	0.000	18:39
Control Test	0.080	18:40
Air Blank	0.000	18:40
Subject Sample # 1	0.131	18:41
Breath Volume	1.894	18:41
Air Blank	0.000	18:41
Air Blank	0.000	18:43
Subject Sample # 2	0.119	18:44
Breath Volume	1.238	18:44
Air Blank	0.000	18:45
Control Test	0.080	18:45
Air Blank	0.000	18:46
Diagnostics Check	OK	18:46

---

Data Download Date/Time: 04/29/2018 01:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARCIA, RAY,  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: KECK, BRANDEN  
ARRESTING OFFICER AGENCY: Oviedo PD  
LAST AGENCY INSPECTION: 03/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: CUTCHENS

DATE/TIME OF TEST: 04/19/2018 04:51:45

---

Diagnostics Check	OK	04:45
Air Blank	0.000	04:45
Control Test	0.081	04:46
Air Blank	0.000	04:46
Subject Sample # 1	0.074	04:47
Breath Volume	3.625	04:47
Air Blank	0.000	04:47
Air Blank	0.000	04:49
Subject Sample # 2	0.073	04:50
Breath Volume	3.687	04:50
Air Blank	0.000	04:50
Control Test	0.080	04:51
Air Blank	0.000	04:51
Diagnostics Check	OK	04:51

---

Data Download Date/Time: 04/29/2018 01:47



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARCIA, RAY,  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: MASSIP, LEONARDO  
ARRESTING OFFICER AGENCY: Oviedo PD  
LAST AGENCY INSPECTION: 03/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: VEAL

DATE/TIME OF TEST: 04/20/2018 22:47:27

---

Diagnostics Check	OK	22:40
Air Blank	0.000	22:41
Control Test	0.080	22:41
Air Blank	0.000	22:42
Subject Sample # 1	0.185	22:42
Breath Volume	1.941	22:42
Air Blank	0.000	22:43
Air Blank	0.000	22:45
Subject Sample # 2	0.181	22:45
Breath Volume	1.824	22:45
Air Blank	0.000	22:46
Control Test	0.079	22:46
Air Blank	0.000	22:47
Diagnostics Check	OK	22:47

---

Data Download Date/Time: 04/29/2018 01:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARCIA, RAY,  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: MERZOYAN, VANEEK  
ARRESTING OFFICER AGENCY: Seminole County SO  
LAST AGENCY INSPECTION: 03/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: ESPY

DATE/TIME OF TEST: 04/21/2018 04:20:23

---

Diagnostics Check	OK	04:13
Air Blank	0.000	04:14
Control Test	0.080	04:14
Air Blank	0.000	04:15
Subject Sample # 1	0.240	04:15
Breath Volume	3.339	04:15
Air Blank	0.000	04:16
Air Blank	0.000	04:18
Subject Sample # 2	0.242	04:18
Breath Volume	3.394	04:18
Air Blank	0.000	04:19
Control Test	0.079	04:19
Air Blank	0.000	04:20
Diagnostics Check	OK	04:20

---

Data Download Date/Time: 04/29/2018 01:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COLBURN,MEGAN,E  
BREATH TEST OPERATOR AGENCY: Casselberry PD  
ARRESTING OFFICER: COLBURN,MEGAN  
ARRESTING OFFICER AGENCY: Casselberry PD  
LAST AGENCY INSPECTION: 03/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 10:21  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 10:55

SUBJECT'S LAST NAME: ZUBLER

DATE/TIME OF TEST: 04/27/2018 11:28:23

---

Diagnostics Check	OK	11:26	
Air Blank	0.000	11:26	
Control Test	0.080	11:27	
Air Blank	0.000	11:27	
		11:27	RFI Detect
Air Blank	0.000	11:28	

---

Data Download Date/Time: 04/29/2018 01:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COLBURN,MEGAN,E  
BREATH TEST OPERATOR AGENCY: Casselberry PD  
ARRESTING OFFICER: COLBURN,MEGAN  
ARRESTING OFFICER AGENCY: Casselberry PD  
LAST AGENCY INSPECTION: 03/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 10:21  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 10:55

SUBJECT'S LAST NAME: ZUBLER

DATE/TIME OF TEST: 04/27/2018 11:36:55

---

Diagnostics Check	OK	11:30
Air Blank	0.000	11:30
Control Test	0.079	11:31
Air Blank	0.000	11:31
Subject Sample # 1	0.000	11:32
Breath Volume	2.125	11:32
Air Blank	0.000	11:32
Air Blank	0.000	11:34
Subject Sample # 2	0.000	11:35
Breath Volume	1.355	11:35
Air Blank	0.000	11:36
Control Test	0.080	11:36
Air Blank	0.000	11:36
Diagnostics Check	OK	11:36

---

Data Download Date/Time: 04/29/2018 01:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ANCION,CALIX,  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: ANCION,CALIX  
ARRESTING OFFICER AGENCY: Seminole County SO  
LAST AGENCY INSPECTION: 03/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 19:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: VERDUGO MORALES

DATE/TIME OF TEST: 04/13/2018 20:36:42

---

Diagnostics Check	OK	20:30
Air Blank	0.000	20:30
Control Test	0.080	20:31
Air Blank	0.000	20:31
Subject Sample # 1	0.222	20:32
Breath Volume	2.335	20:32
Air Blank	0.000	20:32
Air Blank	0.000	20:34
Subject Sample # 2	0.217	20:35
Breath Volume	2.050	20:35
Air Blank	0.000	20:35
Control Test	0.080	20:36
Air Blank	0.000	20:36
Diagnostics Check	OK	20:36

---

Data Download Date/Time: 04/29/2018 01:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: MOREL,ANTHONY  
ARRESTING OFFICER AGENCY: Winter Springs PD  
LAST AGENCY INSPECTION: 03/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:59  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: JEWELL

DATE/TIME OF TEST: 04/14/2018 02:00:10

---

Diagnostics Check	OK	01:49	
Air Blank	0.000	01:50	
Control Test	0.080	01:50	
Air Blank	0.000	01:51	
Subject Sample	# 1 0.105	01:54	Volume Not Met
Breath Volume	0.273	01:54	
Air Blank	0.000	01:54	
Air Blank	0.000	01:56	
Subject Sample	# 2	01:58	Subject Test Refused
Breath Volume	0.230	01:58	
Air Blank	0.000	01:59	
Control Test	0.080	01:59	
Air Blank	0.000	02:00	
Diagnostics Check	OK	02:00	

---

Data Download Date/Time: 04/29/2018 01:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BETHAM,KEITH,D  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: ANCION,CALIX  
ARRESTING OFFICER AGENCY: Seminole County SO  
LAST AGENCY INSPECTION: 03/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 04/14/2018 02:10:55

---

Diagnostics Check	OK	02:04
Air Blank	0.000	02:04
Control Test	0.080	02:05
Air Blank	0.000	02:05
Subject Sample # 1	0.000	02:06
Breath Volume	3.179	02:06
Air Blank	0.000	02:06
Air Blank	0.000	02:08
Subject Sample # 2	0.000	02:09
Breath Volume	3.574	02:09
Air Blank	0.000	02:10
Control Test	0.080	02:10
Air Blank	0.000	02:10
Diagnostics Check	OK	02:10

---

Data Download Date/Time: 04/29/2018 01:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: VOSE,BENJAMIN  
ARRESTING OFFICER AGENCY: Winter Springs PD  
LAST AGENCY INSPECTION: 03/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 04/14/2018 02:56:36

---

Diagnostics Check	OK	02:50
Air Blank	0.000	02:51
Control Test	0.080	02:51
Air Blank	0.000	02:51
Subject Sample # 1	0.157	02:52
Breath Volume	2.335	02:52
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample # 2	0.157	02:55
Breath Volume	2.289	02:55
Air Blank	0.000	02:55
Control Test	0.081	02:56
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

---

Data Download Date/Time: 04/29/2018 01:43



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BETHAM,KEITH,D  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: COPPOLA,TYSON  
ARRESTING OFFICER AGENCY: Longwood PD  
LAST AGENCY INSPECTION: 03/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:03  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 04/14/2018 03:21:54

---

Diagnostics Check		OK	03:10	
Air Blank		0.000	03:10	
Control Test		0.081	03:10	
Air Blank		0.000	03:11	
Subject Sample	# 1	0.184	03:14	Volume Not Met
Breath Volume		0.015	03:14	
Air Blank		0.000	03:15	
Air Blank		0.000	03:17	
Subject Sample	# 2	0.181	03:20	Volume Not Met
Breath Volume		0.109	03:20	
Air Blank		0.000	03:20	
Control Test		0.080	03:21	
Air Blank		0.000	03:21	
Diagnostics Check		OK	03:21	

---

Data Download Date/Time: 04/29/2018 01:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BETHAM,KEITH,D  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: KUIPER,DERRICK  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 03/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:38  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:46

SUBJECT'S LAST NAME: OXLEY

DATE/TIME OF TEST: 04/19/2018 23:39:21

---

Diagnostics Check	OK	23:31
Air Blank	0.000	23:32
Control Test	0.081	23:32
Air Blank	0.000	23:33
Subject Sample # 1	0.188	23:33
Breath Volume	1.441	23:33
Air Blank	0.000	23:34
Air Blank	0.000	23:36
Subject Sample # 2	0.189	23:37
Breath Volume	1.707	23:37
Air Blank	0.000	23:38
Control Test	0.082	23:38
Air Blank	0.000	23:39
Diagnostics Check	OK	23:39

---

Data Download Date/Time: 04/29/2018 01:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ANCION,CALIX,  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: ALBERTS,J  
ARRESTING OFFICER AGENCY: Lake Mary PD  
LAST AGENCY INSPECTION: 03/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 19:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: REINHARDT

DATE/TIME OF TEST: 04/21/2018 20:17:06

---

Diagnostics Check	OK	20:10
Air Blank	0.000	20:11
Control Test	0.080	20:11
Air Blank	0.000	20:12
Subject Sample # 1	0.171	20:12
Breath Volume	3.257	20:12
Air Blank	0.000	20:13
Air Blank	0.000	20:15
Subject Sample # 2	0.169	20:15
Breath Volume	2.820	20:15
Air Blank	0.000	20:16
Control Test	0.079	20:16
Air Blank	0.000	20:17
Diagnostics Check	OK	20:17

---

Data Download Date/Time: 04/29/2018 01:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARCIA, RAY,  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: HALL, JONATHAN  
ARRESTING OFFICER AGENCY: Seminole County SO  
LAST AGENCY INSPECTION: 03/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 05:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: MATEO DOMINGO

DATE/TIME OF TEST: 04/22/2018 06:48:09

---

Diagnostics Check	OK	06:41
Air Blank	0.000	06:41
Control Test	0.081	06:42
Air Blank	0.000	06:42
Subject Sample # 1	0.173	06:43
Breath Volume	2.300	06:43
Air Blank	0.000	06:44
Air Blank	0.000	06:46
Subject Sample # 2	0.171	06:46
Breath Volume	1.773	06:46
Air Blank	0.000	06:47
Control Test	0.080	06:47
Air Blank	0.000	06:48
Diagnostics Check	OK	06:48

---

Data Download Date/Time: 04/29/2018 01:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARCIA, RAY, \  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: MOREL, ANTHONY  
ARRESTING OFFICER AGENCY: Winter Springs PD  
LAST AGENCY INSPECTION: 03/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:23  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: MCMANUS

DATE/TIME OF TEST: 04/25/2018 02:32:13

---

Diagnostics Check	OK	02:29	
Air Blank	0.000	02:29	
Control Test	0.082	02:30	
Air Blank	0.000	02:30	
Subject Sample	# 1	02:30	Subject Test Refused
Breath Volume	0.000	02:30	
Air Blank	0.000	02:31	
Control Test	0.081	02:31	
Air Blank	0.000	02:32	
Diagnostics Check	OK	02:32	

---

Data Download Date/Time: 04/29/2018 01:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARCIA, RAY,  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: PHILLIPS, RON  
ARRESTING OFFICER AGENCY: Casselberry PD  
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: RHOADS

DATE/TIME OF TEST: 04/29/2018 22:48:30

---

Diagnostics Check	OK	22:42
Air Blank	0.000	22:42
Control Test	0.081	22:43
Air Blank	0.000	22:43
Subject Sample # 1	0.146	22:44
Breath Volume	2.976	22:44
Air Blank	0.000	22:44
Air Blank	0.000	22:46
Subject Sample # 2	0.141	22:47
Breath Volume	2.671	22:47
Air Blank	0.000	22:47
Control Test	0.081	22:47
Air Blank	0.000	22:48
Diagnostics Check	OK	22:48

---

Data Download Date/Time: 05/29/2018 00:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARCIA, RAY,  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: HOLIDAY, LUCAS  
ARRESTING OFFICER AGENCY: Maitland PD  
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:32  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: BOOK

DATE/TIME OF TEST: 04/30/2018 04:33:59

---

Diagnostics Check	OK	04:28
Air Blank	0.000	04:28
Control Test	0.080	04:28
Air Blank	0.000	04:29
Subject Sample # 1	0.143	04:29
Breath Volume	1.753	04:29
Air Blank	0.000	04:30
Air Blank	0.000	04:32
Subject Sample # 2	0.139	04:32
Breath Volume	2.000	04:32
Air Blank	0.000	04:33
Control Test	0.081	04:33
Air Blank	0.000	04:33
Diagnostics Check	OK	04:33

---

Data Download Date/Time: 05/29/2018 00:59

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOWATT,NATHAN,E  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: MOWATT,NATHAN  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 03/22/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629802  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 23:23  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: CONLEY

DATE/TIME OF TEST: 04/01/2018 00:07:49

---

Diagnostics Check	OK	00:01
Air Blank	0.000	00:01
Control Test	0.082	00:02
Air Blank	0.000	00:02
Subject Sample # 1	0.089	00:03
Breath Volume	1.582	00:03
Air Blank	0.000	00:03
Air Blank	0.000	00:05
Subject Sample # 2	0.085	00:06
Breath Volume	1.578	00:06
Air Blank	0.000	00:06
Control Test	0.082	00:07
Air Blank	0.000	00:07
Diagnostics Check	OK	00:07

---

Data Download Date/Time: 04/28/2018 02:23



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CORRAO, MICHAEL, W  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: CORRAO, MICHAEL  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 03/22/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629802  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 18:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: CHALKLEY

DATE/TIME OF TEST: 04/01/2018 19:42:25

---

Diagnostics Check	OK	19:36
Air Blank	0.000	19:36
Control Test	0.082	19:36
Air Blank	0.000	19:37
Subject Sample # 1	0.183	19:37
Breath Volume	2.562	19:37
Air Blank	0.000	19:38
Air Blank	0.000	19:40
Subject Sample # 2	0.178	19:40
Breath Volume	1.285	19:40
Air Blank	0.000	19:41
Control Test	0.082	19:41
Air Blank	0.000	19:42
Diagnostics Check	OK	19:42

---

Data Download Date/Time: 04/28/2018 02:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILTSE,DEPUTY WILLIAM,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: WILTSE,DEPUTY WILLIAM  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 03/22/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629802  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: NUNGESTER

DATE/TIME OF TEST: 04/03/2018 02:46:59

---

Diagnostics Check	OK	02:40
Air Blank	0.000	02:40
Control Test	0.082	02:41
Air Blank	0.000	02:41
Subject Sample # 1	0.037	02:42
Breath Volume	2.917	02:42
Air Blank	0.000	02:43
Air Blank	0.000	02:44
Subject Sample # 2	0.037	02:45
Breath Volume	2.687	02:45
Air Blank	0.000	02:46
Control Test	0.082	02:46
Air Blank	0.000	02:46
Diagnostics Check	OK	02:46

---

Data Download Date/Time: 04/28/2018 02:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILTSE,DEPUTY WILLIAM,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: WILTSE,DEPUTY WILLIAM  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 03/22/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629802  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 21:13  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:32

SUBJECT'S LAST NAME: CARO

DATE/TIME OF TEST: 04/06/2018 22:01:03

---

Diagnostics Check	OK	21:54
Air Blank	0.000	21:54
Control Test	0.081	21:55
Air Blank	0.000	21:55
Subject Sample # 1	0.193	21:56
Breath Volume	1.769	21:56
Air Blank	0.000	21:56
Air Blank	0.000	21:58
Subject Sample # 2	0.190	21:59
Breath Volume	2.671	21:59
Air Blank	0.000	22:00
Control Test	0.082	22:00
Air Blank	0.000	22:00
Diagnostics Check	OK	22:01

---

Data Download Date/Time: 04/28/2018 02:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEANE,M,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: VICKERS,J  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 03/22/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629802  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: AQUINO SIERRA

DATE/TIME OF TEST: 04/07/2018 02:05:35

---

Diagnostics Check	OK	02:01	
Air Blank	0.000	02:01	
Control Test	0.082	02:01	
Air Blank	0.000	02:02	
Subject Sample	# 1	02:04	Subject Test Refused
Breath Volume	0.000	02:04	
Air Blank	0.000	02:04	
Control Test	0.081	02:04	
Air Blank	0.000	02:05	
Diagnostics Check	OK	02:05	

---

Data Download Date/Time: 04/28/2018 02:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DORR,RANDALL,S  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: DORR,RANDALL  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 03/22/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629802  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: STATON

DATE/TIME OF TEST: 04/08/2018 02:48:03

---

Diagnostics Check	OK	02:45	
Air Blank	0.000	02:45	
Control Test	0.082	02:45	
Air Blank	0.000	02:46	
Subject Sample	# 1	02:46	Subject Test Refused
Breath Volume	0.000	02:46	
Air Blank	0.000	02:47	
Control Test	0.082	02:47	
Air Blank	0.000	02:47	
Diagnostics Check	OK	02:48	

---

Data Download Date/Time: 04/28/2018 02:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOWATT,NATHAN,E  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: MOWATT,NATHAN  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 03/22/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629802  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: CALLAHAN

DATE/TIME OF TEST: 04/12/2018 01:53:03

---

Diagnostics Check	OK	01:50	
Air Blank	0.000	01:50	
Control Test	0.083	01:50	
Air Blank	0.000	01:51	
Subject Sample	# 1	01:51	Subject Test Refused
Breath Volume	0.000	01:51	
Air Blank	0.000	01:52	
Control Test	0.082	01:52	
Air Blank	0.000	01:52	
Diagnostics Check	OK	01:53	

---

Data Download Date/Time: 04/28/2018 02:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: AMATRUDA,CHRIS  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 03/22/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629802  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: FULLER

DATE/TIME OF TEST: 04/12/2018 03:08:09

---

Diagnostics Check	OK	03:05	
Air Blank	0.000	03:05	
Control Test	0.083	03:06	
Air Blank	0.000	03:06	
Subject Sample	# 1	03:06	Subject Test Refused
Breath Volume	0.000	03:06	
Air Blank	0.000	03:07	
Control Test	0.082	03:07	
Air Blank	0.000	03:08	
Diagnostics Check	OK	03:08	

---

Data Download Date/Time: 04/28/2018 02:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOWATT,NATHAN,E  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: MOWATT,NATHAN  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 03/22/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629802  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 04:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: WHITEHEAD

DATE/TIME OF TEST: 04/13/2018 05:11:09

---

Diagnostics Check		OK	05:02
Air Blank		0.000	05:03
Control Test		0.082	05:03
Air Blank		0.000	05:04
Subject Sample	# 1	0.135	05:04
Breath Volume		1.351	05:04
Air Blank		0.000	05:05
Air Blank		0.000	05:07
Subject Sample	# 2	0.128	05:09
Breath Volume		1.292	05:09
Air Blank		0.000	05:10
Control Test		0.082	05:10
Air Blank		0.000	05:11
Diagnostics Check		OK	05:11

---

Data Download Date/Time: 04/28/2018 02:24



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: AMATRUDA,CHRIS  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 03/22/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629802  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 21:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: WAGNER

DATE/TIME OF TEST: 04/13/2018 22:07:48

---

Diagnostics Check	OK	22:01
Air Blank	0.000	22:02
Control Test	0.082	22:02
Air Blank	0.000	22:02
Subject Sample # 1	0.103	22:03
Breath Volume	1.843	22:03
Air Blank	0.000	22:04
Air Blank	0.000	22:05
Subject Sample # 2	0.101	22:06
Breath Volume	1.718	22:06
Air Blank	0.000	22:06
Control Test	0.082	22:07
Air Blank	0.000	22:07
Diagnostics Check	OK	22:07

---

Data Download Date/Time: 04/28/2018 02:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: AMATRUDA,CHRIS  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 03/22/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629802  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:31  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: GIBBS

DATE/TIME OF TEST: 04/14/2018 03:46:34

---

Diagnostics Check	OK	03:43	
Air Blank	0.000	03:43	
Control Test	0.082	03:44	
Air Blank	0.000	03:44	
Subject Sample	# 1	03:45	Subject Test Refused
Breath Volume	0.000	03:45	
Air Blank	0.000	03:45	
Control Test	0.082	03:45	
Air Blank	0.000	03:46	
Diagnostics Check	OK	03:46	

---

Data Download Date/Time: 04/28/2018 02:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: AMATRUDA,CHRIS  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 03/22/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629802  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 05:43  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 06:05

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 04/14/2018 06:34:36

---

Diagnostics Check		OK	06:27
Air Blank		0.000	06:28
Control Test		0.082	06:28
Air Blank		0.000	06:28
Subject Sample	# 1	0.206	06:30
Breath Volume		2.570	06:30
Air Blank		0.000	06:30
Air Blank		0.000	06:32
Subject Sample	# 2	0.207	06:32
Breath Volume		2.375	06:32
Air Blank		0.000	06:33
Control Test		0.081	06:33
Air Blank		0.000	06:34
Diagnostics Check		OK	06:34

---

Data Download Date/Time: 04/28/2018 02:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: AMATRUDA,CHRIS  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 03/22/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629802  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 21:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:09

SUBJECT'S LAST NAME: MAWI

DATE/TIME OF TEST: 04/14/2018 22:39:25

---

Diagnostics Check	OK	22:32
Air Blank	0.000	22:33
Control Test	0.081	22:33
Air Blank	0.000	22:34
Subject Sample # 1	0.246	22:34
Breath Volume	1.812	22:34
Air Blank	0.000	22:35
Air Blank	0.000	22:37
Subject Sample # 2	0.253	22:37
Breath Volume	1.445	22:37
Air Blank	0.000	22:38
Control Test	0.081	22:38
Air Blank	0.000	22:39
Diagnostics Check	OK	22:39

---

Data Download Date/Time: 04/28/2018 02:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: AMATRUDA,CHRIS  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 03/22/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629802  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 00:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: DOTY

DATE/TIME OF TEST: 04/15/2018 01:29:11

---

Diagnostics Check	OK	01:22
Air Blank	0.000	01:22
Control Test	0.082	01:23
Air Blank	0.000	01:23
Subject Sample # 1	0.000	01:24
Breath Volume	2.855	01:24
Air Blank	0.000	01:24
Air Blank	0.000	01:26
Subject Sample # 2	0.000	01:27
Breath Volume	2.222	01:27
Air Blank	0.000	01:28
Control Test	0.082	01:28
Air Blank	0.000	01:29
Diagnostics Check	OK	01:29

---

Data Download Date/Time: 04/28/2018 02:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REFUSE, RONY,  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: REFUSE, RONY  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 03/22/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629802  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 19:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:53

SUBJECT'S LAST NAME: SOUTHERN

DATE/TIME OF TEST: 04/15/2018 21:20:09

---

Diagnostics Check	OK	21:13
Air Blank	0.000	21:13
Control Test	0.082	21:13
Air Blank	0.000	21:14
Subject Sample # 1	0.232	21:14
Breath Volume	1.464	21:14
Air Blank	0.000	21:15
Air Blank	0.000	21:17
Subject Sample # 2	0.225	21:18
Breath Volume	1.949	21:18
Air Blank	0.000	21:19
Control Test	0.082	21:19
Air Blank	0.000	21:20
Diagnostics Check	OK	21:20

---

Data Download Date/Time: 04/28/2018 02:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILTSE,DEPUTY W.,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: LONGMIRE,TROOPER B.  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 03/22/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629802  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 18:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:01

SUBJECT'S LAST NAME: MADSON

DATE/TIME OF TEST: 04/16/2018 19:28:48

---

Diagnostics Check	OK	19:21
Air Blank	0.000	19:22
Control Test	0.082	19:22
Air Blank	0.000	19:23
Subject Sample # 1	0.245	19:24
Breath Volume	1.371	19:24
Air Blank	0.000	19:24
Air Blank	0.000	19:26
Subject Sample # 2	0.265	19:27
Breath Volume	1.640	19:27
Air Blank	0.000	19:27
Control Test	0.080	19:28
Air Blank	0.000	19:28
Diagnostics Check	OK	19:28

---

Data Download Date/Time: 04/28/2018 02:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILTSE,DEPUTY W.,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: WILTSE,DEPUTY W.  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 03/22/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629802  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 00:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: NGUYEN

DATE/TIME OF TEST: 04/17/2018 01:43:37

---

Diagnostics Check	OK	01:36
Air Blank	0.000	01:37
Control Test	0.082	01:37
Air Blank	0.000	01:37
Subject Sample # 1	0.081	01:39
Breath Volume	1.218	01:39
Air Blank	0.000	01:39
Air Blank	0.000	01:41
Subject Sample # 2	0.085	01:42
Breath Volume	1.203	01:42
Air Blank	0.000	01:42
Control Test	0.083	01:43
Air Blank	0.000	01:43
Diagnostics Check	OK	01:43

---

Data Download Date/Time: 04/28/2018 02:24



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILTSE,DEPUTY WILLIAM,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: MARTINEZ,OFFICER A.  
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)  
LAST AGENCY INSPECTION: 03/22/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629802  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 17:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:01

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 04/17/2018 18:28:40

---

Diagnostics Check	OK	18:22
Air Blank	0.000	18:22
Control Test	0.082	18:22
Air Blank	0.000	18:23
Subject Sample # 1	0.102	18:24
Breath Volume	2.988	18:24
Air Blank	0.000	18:24
Air Blank	0.000	18:26
Subject Sample # 2	0.095	18:27
Breath Volume	1.503	18:27
Air Blank	0.000	18:27
Control Test	0.083	18:28
Air Blank	0.000	18:28
Diagnostics Check	OK	18:28

---

Data Download Date/Time: 04/28/2018 02:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILTSE,DEPUTY WILLIAM,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: WILTSE,DEPUTY WILLIAM  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 03/22/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629802  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 20:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: JASTROCH

DATE/TIME OF TEST: 04/17/2018 21:21:06

---

Diagnostics Check	OK	21:12
Air Blank	0.000	21:13
Control Test	0.082	21:13
Air Blank	0.000	21:14
Subject Sample # 1	0.219	21:16
Breath Volume	2.281	21:16
Air Blank	0.000	21:17
Air Blank	0.000	21:18
Subject Sample # 2	0.219	21:19
Breath Volume	2.734	21:19
Air Blank	0.000	21:20
Control Test	0.081	21:20
Air Blank	0.000	21:21
Diagnostics Check	OK	21:21

---

Data Download Date/Time: 04/28/2018 02:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: AMATRUDA,CHRIS  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 03/22/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629802  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 21:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: BARRY

DATE/TIME OF TEST: 04/18/2018 22:45:47

---

Diagnostics Check	OK	22:38
Air Blank	0.000	22:39
Control Test	0.082	22:39
Air Blank	0.000	22:40
Subject Sample # 1	0.100	22:40
Breath Volume	1.136	22:40
Air Blank	0.000	22:41
Air Blank	0.000	22:43
Subject Sample # 2	0.092	22:44
Breath Volume	1.109	22:44
Air Blank	0.000	22:44
Control Test	0.082	22:45
Air Blank	0.000	22:45
Diagnostics Check	OK	22:45

---

Data Download Date/Time: 04/28/2018 02:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: AMATRUDA,CHRIS  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 03/22/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629802  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 04:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:57

SUBJECT'S LAST NAME: AMATRUDA

DATE/TIME OF TEST: 04/20/2018 05:22:04

---

Diagnostics Check	OK	05:19	
Air Blank	0.000	05:19	
Control Test	0.083	05:20	
Air Blank	0.000	05:20	
Subject Sample	# 1	05:20	Subject Test Refused
Breath Volume	0.000	05:20	
Air Blank	0.000	05:21	
Control Test	0.082	05:21	
Air Blank	0.000	05:21	
Diagnostics Check	OK	05:22	

---

Data Download Date/Time: 04/28/2018 02:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILTSE,DEPUTY WILLIAM,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: WILTSE,DEPUTY WILLIAM  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 03/22/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629802  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 20:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 04/20/2018 21:15:50

---

Diagnostics Check	OK	21:13	
Air Blank	0.000	21:13	
Control Test	0.082	21:13	
Air Blank	0.000	21:14	
Subject Sample	# 1	21:14	Subject Test Refused
Breath Volume	0.000	21:14	
Air Blank	0.000	21:14	
Control Test	0.083	21:15	
Air Blank	0.000	21:15	
Diagnostics Check	OK	21:15	

---

Data Download Date/Time: 04/28/2018 02:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILTSE,DEPUTY WILLIAM,  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: WILTSE,DEPUTY WILLIAM  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 03/22/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629802  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 05:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:10

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 04/21/2018 06:33:04

---

Diagnostics Check	OK	06:30	
Air Blank	0.000	06:30	
Control Test	0.083	06:31	
Air Blank	0.000	06:31	
Subject Sample	# 1	06:31	Subject Test Refused
Breath Volume	0.000	06:31	
Air Blank	0.000	06:32	
Control Test	0.082	06:32	
Air Blank	0.000	06:32	
Diagnostics Check	OK	06:33	

---

Data Download Date/Time: 04/28/2018 02:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: AMATRUDA,CHRIS  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 03/22/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629802  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: MEYER

DATE/TIME OF TEST: 04/22/2018 02:20:50

---

Diagnostics Check	OK	02:14
Air Blank	0.000	02:14
Control Test	0.082	02:15
Air Blank	0.000	02:15
Subject Sample # 1	0.141	02:16
Breath Volume	1.457	02:16
Air Blank	0.000	02:16
Air Blank	0.000	02:18
Subject Sample # 2	0.136	02:19
Breath Volume	1.503	02:19
Air Blank	0.000	02:19
Control Test	0.082	02:20
Air Blank	0.000	02:20
Diagnostics Check	OK	02:20

---

Data Download Date/Time: 04/28/2018 02:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNAVELY,KENT,J  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: SNAVELY,KENT  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 03/22/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629802  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 16:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:58

SUBJECT'S LAST NAME: JOHNS

DATE/TIME OF TEST: 04/22/2018 18:31:03

---

Diagnostics Check	OK	18:24
Air Blank	0.000	18:24
Control Test	0.083	18:24
Air Blank	0.000	18:25
Subject Sample # 1	0.024	18:26
Breath Volume	1.957	18:26
Air Blank	0.000	18:26
Air Blank	0.000	18:28
Subject Sample # 2	0.021	18:29
Breath Volume	1.281	18:29
Air Blank	0.000	18:30
Control Test	0.082	18:30
Air Blank	0.000	18:30
Diagnostics Check	OK	18:31

---

Data Download Date/Time: 04/28/2018 02:24



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILTSE,DEPUTY WILLIAM,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: WILTSE,DEPUTY WILLIAM  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 03/22/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629802  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 21:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: PAES

DATE/TIME OF TEST: 04/22/2018 22:01:38

---

Diagnostics Check	OK	21:58	
Air Blank	0.000	21:59	
Control Test	0.082	21:59	
Air Blank	0.000	22:00	
Subject Sample	# 1	22:00	Subject Test Refused
Breath Volume	0.000	22:00	
Air Blank	0.000	22:00	
Control Test	0.082	22:01	
Air Blank	0.000	22:01	
Diagnostics Check	OK	22:01	

---

Data Download Date/Time: 04/28/2018 02:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LONGMIRE,BENJAMIN,E  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: LONGMIRE,BENJAMIN  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 03/22/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629802  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 10:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 04/22/2018 23:44:38

---

Diagnostics Check	OK	23:38
Air Blank	0.000	23:38
Control Test	0.081	23:38
Air Blank	0.000	23:39
Subject Sample # 1	0.212	23:39
Breath Volume	2.519	23:39
Air Blank	0.000	23:40
Air Blank	0.000	23:42
Subject Sample # 2	0.216	23:42
Breath Volume	3.453	23:42
Air Blank	0.000	23:43
Control Test	0.084	23:44
Air Blank	0.000	23:44
Diagnostics Check	OK	23:44

---

Data Download Date/Time: 04/28/2018 02:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOWATT,NATHAN,E  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: MOWATT,NATHAN  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 03/22/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629802  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: CARD

DATE/TIME OF TEST: 04/26/2018 03:04:48

---

Diagnostics Check	OK	03:01	
Air Blank	0.000	03:01	
Control Test	0.082	03:02	
Air Blank	0.000	03:02	
Subject Sample	# 1	03:03	Subject Test Refused
Breath Volume	0.000	03:03	
Air Blank	0.000	03:03	
Control Test	0.082	03:04	
Air Blank	0.000	03:04	
Diagnostics Check	OK	03:04	

---

Data Download Date/Time: 04/28/2018 02:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: AMATRUDA,CHRIS  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 03/22/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629802  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:01  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: SIBER

DATE/TIME OF TEST: 04/27/2018 01:52:09

---

Diagnostics Check	OK	01:49	
Air Blank	0.000	01:49	
Control Test	0.082	01:50	
Air Blank	0.000	01:50	
Subject Sample	# 1	01:50	Subject Test Refused
Breath Volume	0.000	01:50	
Air Blank	0.000	01:51	
Control Test	0.082	01:51	
Air Blank	0.000	01:52	
Diagnostics Check	OK	01:52	

---

Data Download Date/Time: 04/28/2018 02:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: AMATRUDA,CHRIS  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 04/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629802  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: OWEN

DATE/TIME OF TEST: 04/28/2018 02:48:14

---

Diagnostics Check	OK	02:45	
Air Blank	0.000	02:45	
Control Test	0.082	02:46	
Air Blank	0.000	02:46	
Subject Sample	# 1	02:46	Subject Test Refused
Breath Volume	0.000	02:46	
Air Blank	0.000	02:47	
Control Test	0.081	02:47	
Air Blank	0.000	02:48	
Diagnostics Check	OK	02:48	

---

Data Download Date/Time: 05/31/2018 01:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: AMATRUDA,CHRIS  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 04/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629802  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 22:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: AVILES

DATE/TIME OF TEST: 04/28/2018 23:59:07

---

Diagnostics Check		OK	23:45	
Air Blank		0.000	23:45	
Control Test		0.082	23:45	
Air Blank		0.000	23:46	
Subject Sample	# 1	0.208	23:49	Volume Not Met
Breath Volume		0.296	23:49	
Air Blank		0.000	23:50	
Air Blank		0.000	23:51	
Subject Sample	# 2	0.244	23:53	
Breath Volume		1.203	23:53	
Air Blank		0.000	23:54	
Air Blank		0.000	23:56	
Subject Sample	# 3	0.237	23:57	
Breath Volume		1.402	23:57	
Air Blank		0.000	23:58	
Control Test		0.080	23:58	
Air Blank		0.000	23:59	
Diagnostics Check		OK	23:59	

---

Data Download Date/Time: 05/31/2018 01:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REEVES,MARK,O  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: NETTLES,JIMMY  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 03/22/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629802  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 11:01  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 11:48

SUBJECT'S LAST NAME: PETROSKI

DATE/TIME OF TEST: 04/06/2018 12:14:48

---

Diagnostics Check	OK	12:08
Air Blank	0.000	12:08
Control Test	0.080	12:09
Air Blank	0.000	12:09
Subject Sample # 1	0.328	12:10
Breath Volume	3.339	12:10
Air Blank	0.000	12:10
Air Blank	0.000	12:12
Subject Sample # 2	0.320	12:13
Breath Volume	2.265	12:13
Air Blank	0.000	12:13
Control Test	0.079	12:14
Air Blank	0.000	12:14
Diagnostics Check	OK	12:14

---

Data Download Date/Time: 04/27/2018 22:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSON,BRUCE,M  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: COOK,CHRIS  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 03/22/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG728303  
DRY GAS EXPIRY DATE: 10/10/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: OLENDER

DATE/TIME OF TEST: 04/09/2018 01:20:14

---

Diagnostics Check	OK	01:17	
Air Blank	0.000	01:17	
Control Test	0.080	01:18	
Air Blank	0.000	01:18	
Subject Sample	# 1	01:18	Subject Test Refused
Breath Volume	0.000	01:18	
Air Blank	0.000	01:19	
Control Test	0.080	01:19	
Air Blank	0.000	01:20	
Diagnostics Check	OK	01:20	

---

Data Download Date/Time: 04/27/2018 22:21



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSON, BRUCE, M  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: CORSER, W  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 03/22/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG728303  
DRY GAS EXPIRY DATE: 10/10/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: GUTIERREZ AREVALO

DATE/TIME OF TEST: 04/09/2018 02:06:35

---

Diagnostics Check	OK	01:57
Air Blank	0.000	01:57
Control Test	0.081	01:57
Air Blank	0.000	01:58
Subject Sample # 1	0.197	01:58
Breath Volume	3.082	01:58
Air Blank	0.000	01:59
Air Blank	0.000	02:01
Subject Sample # 2	0.176	02:01
Breath Volume	2.171	02:01
Air Blank	0.000	02:02
Air Blank	0.000	02:04
Subject Sample # 3	0.197	02:05
Breath Volume	2.757	02:05
Air Blank	0.000	02:05
Control Test	0.080	02:05
Air Blank	0.000	02:06
Diagnostics Check	OK	02:06

---

Data Download Date/Time: 04/27/2018 22:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SZYMANSKI,JACEK,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: FEIDT,EDWARD  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 03/22/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG728303  
DRY GAS EXPIRY DATE: 10/10/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: RUIZ-CABAN

DATE/TIME OF TEST: 04/10/2018 23:13:54

---

Diagnostics Check	OK	23:07
Air Blank	0.000	23:08
Control Test	0.081	23:08
Air Blank	0.000	23:09
Subject Sample # 1	0.110	23:09
Breath Volume	2.609	23:09
Air Blank	0.000	23:10
Air Blank	0.000	23:12
Subject Sample # 2	0.109	23:12
Breath Volume	3.234	23:12
Air Blank	0.000	23:13
Control Test	0.080	23:13
Air Blank	0.000	23:13
Diagnostics Check	OK	23:13

---

Data Download Date/Time: 04/27/2018 22:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALLAIRE,JESSICA,L  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: MALDONADO,J  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 03/22/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG728303  
DRY GAS EXPIRY DATE: 10/10/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: BOWEN

DATE/TIME OF TEST: 04/18/2018 04:24:23

---

Diagnostics Check	OK	04:18
Air Blank	0.000	04:18
Control Test	0.080	04:18
Air Blank	0.000	04:19
Subject Sample # 1	0.104	04:19
Breath Volume	3.351	04:19
Air Blank	0.000	04:20
Air Blank	0.000	04:22
Subject Sample # 2	0.105	04:22
Breath Volume	3.363	04:22
Air Blank	0.000	04:23
Control Test	0.080	04:23
Air Blank	0.000	04:24
Diagnostics Check	OK	04:24

---

Data Download Date/Time: 04/27/2018 22:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KASELAK,STEPHEN,G  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: RICOTTILLI,JOSHUA  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 03/22/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG728303  
DRY GAS EXPIRY DATE: 10/10/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:53  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: ERNST

DATE/TIME OF TEST: 04/21/2018 03:26:04

---

Diagnostics Check	OK	03:23	
Air Blank	0.000	03:23	
Control Test	0.080	03:24	
Air Blank	0.000	03:24	
Subject Sample	# 1	03:24	Subject Test Refused
Breath Volume	0.000	03:24	
Air Blank	0.000	03:25	
Control Test	0.079	03:25	
Air Blank	0.000	03:25	
Diagnostics Check	OK	03:26	

---

Data Download Date/Time: 04/27/2018 22:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALLAIRE,JESSICA,L  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: LILLEY,KATHLEEN  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 03/22/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG728303  
DRY GAS EXPIRY DATE: 10/10/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: DIEHL

DATE/TIME OF TEST: 04/23/2018 00:03:39

---

Diagnostics Check	OK	00:00	
Air Blank	0.000	00:01	
Control Test	0.080	00:01	
Air Blank	0.000	00:01	
Subject Sample	# 1	00:02	Subject Test Refused
Breath Volume	0.000	00:02	
Air Blank	0.000	00:02	
Control Test	0.081	00:03	
Air Blank	0.000	00:03	
Diagnostics Check	OK	00:03	

---

Data Download Date/Time: 04/27/2018 22:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALLAIRE,JESSICA,L  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: PUPKE,LUKE  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 03/22/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG728303  
DRY GAS EXPIRY DATE: 10/10/2019  
VIOLATION CODE: DUI  
ARREST TIME: 04:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:46

SUBJECT'S LAST NAME: ROMAN

DATE/TIME OF TEST: 04/23/2018 05:10:14

---

Diagnostics Check	OK	05:07	
Air Blank	0.000	05:07	
Control Test	0.080	05:08	
Air Blank	0.000	05:08	
Subject Sample	# 1	05:08	Subject Test Refused
Breath Volume	0.000	05:08	
Air Blank	0.000	05:09	
Control Test	0.082	05:09	
Air Blank	0.000	05:10	
Diagnostics Check	OK	05:10	

---

Data Download Date/Time: 04/27/2018 22:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHRISTIE,MATT,J  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: TEMPLE,LIZ  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 03/22/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG728303  
DRY GAS EXPIRY DATE: 10/10/2019  
VIOLATION CODE: DUI  
ARREST TIME: 04:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: ALLEVA

DATE/TIME OF TEST: 04/25/2018 05:29:55

---

Diagnostics Check	OK	05:26	
Air Blank	0.000	05:26	
Control Test	0.081	05:27	
Air Blank	0.000	05:27	
Subject Sample	# 1	05:28	Subject Test Refused
Breath Volume	0.000	05:28	
Air Blank	0.000	05:29	
Control Test	0.076	05:29	
Air Blank	0.000	05:29	
Diagnostics Check	OK	05:29	

---

Data Download Date/Time: 04/27/2018 22:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEANE,M,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: RICOTTILLI,J  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 03/22/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG728303  
DRY GAS EXPIRY DATE: 10/10/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: HEANEY

DATE/TIME OF TEST: 04/26/2018 00:33:55

---

Diagnostics Check		OK	00:25
Air Blank		0.000	00:25
Control Test		0.081	00:26
Air Blank		0.000	00:26
Subject Sample	# 1	0.213	00:28
Breath Volume		2.476	00:28
Air Blank		0.000	00:29
Air Blank		0.000	00:30
Subject Sample	# 2	0.210	00:32
Breath Volume		2.156	00:32
Air Blank		0.000	00:33
Control Test		0.078	00:33
Air Blank		0.000	00:33
Diagnostics Check		OK	00:33

---

Data Download Date/Time: 04/27/2018 22:21