Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: MORRIS,ERIC
ARRESTING OFFICER AGENCY: Winter Springs PD
LAST AGENCY INSPECTION: 04/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: WIECHERT

DATE/TIME OF TEST: 04/29/2018 04:27:00

Diagnostics Check		OK	04:20
Air Blank		0.000	04:20
Control Test		0.078	04:20
Air Blank		0.000	04:21
Subject Sample	# 1	0.181	04:22
Breath Volume		2.207	04:22
Air Blank		0.000	04:22
Air Blank		0.000	04:24
Subject Sample	# 2	0.178	04:25
Breath Volume		3.015	04:25
Air Blank		0.000	04:26
Control Test		0.079	04:26
Air Blank		0.000	04:26
Diagnostics Check		OK	04:27

Data Download Date/Time: 05/29/2018 01:05

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001000

INSTRUMENT REGISTERED TO: COCOA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: HALL,NICHOLAS
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 03/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: ARNOLD

DATE/TIME OF TEST: 04/11/2018 01:13:24

Diagnostics Check		OK	01:10	
Air Blank		0.000	01:10	
Control Test		0.079	01:10	
Air Blank		0.000	01:11	
Subject Sample	# 1		01:12	Subject Test Refused
Breath Volume		0.000	01:12	
Air Blank		0.000	01:12	
Control Test		0.079	01:12	
Air Blank		0.000	01:13	
Diagnostics Check		OK	01:13	

Data Download Date/Time: 04/26/2018 22:05

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001000

INSTRUMENT REGISTERED TO: COCOA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: HALL,NICHOLAS
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 03/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: CHIOMA

DATE/TIME OF TEST: 04/15/2018 22:02:52

Diagnostics Check		OK	21:55
Air Blank		0.000	21:56
Control Test		0.078	21:56
Air Blank		0.000	21:57
Subject Sample	# 1	0.169	21:58
Breath Volume		3.191	21:58
Air Blank		0.000	21:59
Air Blank		0.000	22:00
Subject Sample	# 2	0.171	22:01
Breath Volume		1.960	22:01
Air Blank		0.000	22:01
Control Test		0.077	22:02
Air Blank		0.000	22:02
Diagnostics Check		OK	22:02

Data Download Date/Time: 04/26/2018 22:05

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001000

INSTRUMENT REGISTERED TO: COCOA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: HALL,NICHOLAS
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 03/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: RAY

DATE/TIME OF TEST: 04/23/2018 03:00:11

Diagnostics Check		OK	02:57	
Air Blank		0.000	02:57	
Control Test		0.079	02:58	
Air Blank		0.000	02:58	
Subject Sample	# 1		02:58	Subject Test Refused
Breath Volume		0.000	02:58	
Air Blank		0.000	02:59	
Control Test		0.079	02:59	
Air Blank		0.000	03:00	
Diagnostics Check		OK	03:00	

Data Download Date/Time: 04/26/2018 22:05

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOUSMAN,JOHN,D
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: HOUSMAN,JOHN
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 03/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 14:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:33

SUBJECT'S LAST NAME: THACKER

DATE/TIME OF TEST: 04/05/2018 15:03:26

Diagnostics Check		OK	14:56
Air Blank		0.000	14:57
Control Test		0.078	14:57
Air Blank		0.000	14:58
Subject Sample	# 1	0.293	14:58
Breath Volume		2.480	14:58
Air Blank		0.000	14:59
Air Blank		0.000	15:01
Subject Sample	# 2	0.279	15:01
Breath Volume		2.472	15:01
Air Blank		0.000	15:02
Control Test		0.078	15:02
Air Blank		0.000	15:03
Diagnostics Check		OK	15:03

Data Download Date/Time: 04/28/2018 06:20

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIEBELL,DON,R
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: TRIEBELL,DON
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 03/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: BARNETT

DATE/TIME OF TEST: 04/20/2018 01:20:42

Diagnostics Check		OK	01:13
Air Blank		0.000	01:13
Control Test		0.077	01:14
Air Blank		0.000	01:14
Subject Sample	# 1	0.229	01:15
Breath Volume		1.382	01:15
Air Blank		0.000	01:16
Air Blank		0.000	01:18
Subject Sample	# 2	0.223	01:19
Breath Volume		1.523	01:19
Air Blank		0.000	01:19
Control Test		0.076	01:20
Air Blank		0.000	01:20
Diagnostics Check		OK	01:20

Data Download Date/Time: 04/28/2018 06:20

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCRANN,SEAN,M
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: MCCRANN,SEAN
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 04/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 04/29/2018 01:19:59

Diagnostics Check		OK	01:10
Air Blank		0.000	01:11
Control Test		0.077	01:11
Air Blank		0.000	01:11
Subject Sample	# 1	0.202	01:14
Breath Volume		1.226	01:14
Air Blank		0.000	01:14
Air Blank		0.000	01:16
Subject Sample	# 2	0.199	01:18
Breath Volume		2.394	01:18
Air Blank		0.000	01:19
Control Test		0.077	01:19
Air Blank		0.000	01:19
Diagnostics Check		OK	01:19

Data Download Date/Time: 05/25/2018 14:08

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUSE,RONNIE,L
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: GICKER,CHRISTOPHER
ARRESTING OFFICER AGENCY: Wauchula PD
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 04/28/2018 04:03:52

Diagnostics Check		OK	03:54
Air Blank		0.000	03:54
Control Test		0.079	03:55
Air Blank		0.000	03:55
Subject Sample	# 1	0.149	03:58
Breath Volume		2.671	03:58
Air Blank		0.000	03:58
Air Blank		0.000	04:00
Subject Sample	# 2	0.131	04:02
Breath Volume		1.964	04:02
Air Blank		0.000	04:02
Control Test		0.079	04:03
Air Blank		0.000	04:03
Diagnostics Check		OK	04:03

Data Download Date/Time: 05/30/2018 06:15

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUSE,RONNIE,L
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: WARD,SHANE
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 18:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:18

SUBJECT'S LAST NAME: ZURITA

DATE/TIME OF TEST: 04/28/2018 19:45:44

Diagnostics Check		OK	19:39	
Air Blank		0.000	19:39	
Control Test		0.079	19:39	
Air Blank		0.000	19:40	
Subject Sample	# 1	0.369	19:41	
Breath Volume		2.191	19:41	
Air Blank		0.000	19:41	
Air Blank		0.000	19:43	
Subject Sample	# 2		19:44	Slope Not Met
Breath Volume		0.683	19:44	
Air Blank		0.000	19:44	
Control Test		0.079	19:45	
Air Blank		0.000	19:45	
Diagnostics Check		OK	19:45	

Data Download Date/Time: 05/30/2018 06:15

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUSE,RONNIE,L
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: WARD,SHANE
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 18:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:18

SUBJECT'S LAST NAME: ZURITA

DATE/TIME OF TEST: 04/28/2018 19:55:04

Diagnostics Check		OK	19:48
Air Blank		0.000	19:48
Control Test		0.079	19:49
Air Blank		0.000	19:49
Subject Sample	# 1	0.368	19:50
Breath Volume		2.601	19:50
Air Blank		0.000	19:51
Air Blank		0.000	19:53
Subject Sample	# 2	0.367	19:53
Breath Volume		2.128	19:53
Air Blank		0.000	19:54
Control Test		0.078	19:54
Air Blank		0.000	19:54
Diagnostics Check		OK	19:55

Data Download Date/Time: 05/30/2018 06:15

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

INSTRUMENT REGISTERED TO: INDIALANTIC P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOVEL,RYAN,M
BREATH TEST OPERATOR AGENCY: Indialantic PD
ARRESTING OFFICER: CRONIN,SEAN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 04/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 04/21/2018 23:41:29

Diagnostics Check		OK	23:35
Air Blank		0.000	23:35
Control Test		0.080	23:35
Air Blank		0.000	23:36
Subject Sample	# 1	0.138	23:36
Breath Volume		2.257	23:36
Air Blank		0.000	23:37
Air Blank		0.000	23:39
Subject Sample	# 2	0.138	23:39
Breath Volume		2.734	23:39
Air Blank		0.000	23:40
Control Test		0.080	23:40
Air Blank		0.000	23:41
Diagnostics Check		OK	23:41

Data Download Date/Time: 05/25/2018 07:46

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

INSTRUMENT REGISTERED TO: INDIALANTIC P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOVEL,RYAN,M
BREATH TEST OPERATOR AGENCY: Indialantic PD
ARRESTING OFFICER: DOVEL,RYAN
ARRESTING OFFICER AGENCY: Indialantic PD
LAST AGENCY INSPECTION: 04/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 04/26/2018 02:44:50

Diagnostics Check		OK	02:31	
Air Blank		0.000	02:31	
Control Test		0.080	02:32	
Air Blank		0.000	02:32	
Subject Sample	# 1	0.000	02:33	
Breath Volume		1.371	02:33	
Air Blank		0.000	02:33	
Air Blank		0.000	02:35	
Subject Sample	# 2		02:39	No Sample Provided
Breath Volume		0.000	02:39	
Air Blank		0.000	02:39	
Air Blank		0.000	02:41	
Subject Sample	# 3	0.000	02:43	
Breath Volume		1.582	02:43	
Air Blank		0.000	02:43	
Control Test		0.081	02:44	
Air Blank		0.000	02:44	
Diagnostics Check		OK	02:44	

Data Download Date/Time: 05/25/2018 07:46

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKSON,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: JACKON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 04/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 16:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:20

SUBJECT'S LAST NAME: SINDO

DATE/TIME OF TEST: 04/16/2018 16:52:51

Diagnostics Check		OK	16:46
Air Blank		0.000	16:47
Control Test		0.082	16:47
Air Blank		0.000	16:48
Subject Sample	# 1	0.000	16:48
Breath Volume		2.171	16:48
Air Blank		0.000	16:49
Air Blank		0.000	16:51
Subject Sample	# 2	0.000	16:51
Breath Volume		2.687	16:51
Air Blank		0.000	16:51
Control Test		0.082	16:52
Air Blank		0.000	16:52
Diagnostics Check		OK	16:52

Data Download Date/Time: 05/13/2018 00:57

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCKER,WILLIAM,
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: TUCKER,WILLIAM
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 04/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: MARCIA VELASQUEZ

DATE/TIME OF TEST: 04/22/2018 00:20:31

Diagnostics Check	OK	00:18	
Air Blank	0.000	00:19	
Control Test	0.082	00:19	
Air Blank		00:20	RFI Detect
Air Blank	0.000	00:20	

Data Download Date/Time: 05/13/2018 00:57

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCKER,WILLIAM,
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: TUCKER,WILLIAM
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 04/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: MARCIA VELASQUEZ

DATE/TIME OF TEST: 04/22/2018 00:32:52

Diagnostics Check		OK	00:23
Air Blank		0.000	00:24
Control Test		0.082	00:24
Air Blank		0.000	00:24
Subject Sample	# 1	0.187	00:25
Breath Volume		2.851	00:25
Air Blank		0.000	00:26
Air Blank		0.000	00:27
Subject Sample	# 2	0.164	00:28
Breath Volume		1.539	00:28
Air Blank		0.000	00:29
Air Blank		0.000	00:30
Subject Sample	# 3	0.181	00:31
Breath Volume		3.214	00:31
Air Blank		0.000	00:31
Control Test		0.079	00:32
Air Blank		0.000	00:32
Diagnostics Check		OK	00:32

Data Download Date/Time: 05/13/2018 00:57

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,LATORIAN,T
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: TURNER,LATORIAN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 04/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 18:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:13

SUBJECT'S LAST NAME: PREVALUS

DATE/TIME OF TEST: 04/24/2018 20:16:54

Diagnostics Check		OK	20:14	
Air Blank		0.000	20:14	
Control Test		0.083	20:14	
Air Blank		0.000	20:15	
Subject Sample	# 1		20:15	Subject Test Refused
Breath Volume		0.000	20:15	
Air Blank		0.000	20:15	
Control Test		0.082	20:16	
Air Blank		0.000	20:16	
Diagnostics Check		OK	20:16	

Data Download Date/Time: 05/13/2018 00:57

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,RAY,M
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: HOLMES,KALEN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 04/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 17:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: SCARBOROUGH

DATE/TIME OF TEST: 04/29/2018 17:52:03

Diagnostics Check		OK	17:49	
Air Blank		0.000	17:49	
Control Test		0.082	17:49	
Air Blank		0.000	17:50	
Subject Sample	# 1		17:50	Subject Test Refused
Breath Volume		0.000	17:50	
Air Blank		0.000	17:51	
Control Test		0.082	17:51	
Air Blank		0.000	17:51	
Diagnostics Check		OK	17:52	

Data Download Date/Time: 05/13/2018 00:57

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,B,
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: DANTI,T
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 03/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: SIERRA

DATE/TIME OF TEST: 04/13/2018 03:25:49

Diagnostics Check		OK	03:19
Air Blank		0.000	03:20
Control Test		0.080	03:20
Air Blank		0.000	03:20
Subject Sample	# 1	0.268	03:21
Breath Volume		2.296	03:21
Air Blank		0.000	03:21
Air Blank		0.000	03:23
Subject Sample	# 2	0.254	03:24
Breath Volume		1.742	03:24
Air Blank		0.000	03:24
Control Test		0.080	03:25
Air Blank		0.000	03:25
Diagnostics Check		OK	03:25

Data Download Date/Time: 04/25/2018 20:19

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROE,ROBERT,D
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: MEE,TIMOTHY
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 04/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: SEBRING

DATE/TIME OF TEST: 04/28/2018 01:53:04

Diagnostics Check		OK	01:49	
Air Blank		0.000	01:50	
Control Test		0.079	01:50	
Air Blank		0.000	01:50	
Subject Sample	# 1		01:51	Subject Test Refused
Breath Volume		0.000	01:51	
Air Blank		0.000	01:52	
Control Test		0.079	01:52	
Air Blank		0.000	01:52	
Diagnostics Check		OK	01:53	

Data Download Date/Time: 05/27/2018 16:03

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROE,ROBERT,D
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SALVADOR,ELDER
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 04/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: CARSON

DATE/TIME OF TEST: 04/29/2018 01:13:41

Diagnostics Check		OK	01:10	
Air Blank		0.000	01:10	
Control Test		0.079	01:10	
Air Blank		0.000	01:11	
Subject Sample	# 1		01:12	Subject Test Refused
Breath Volume		0.000	01:12	
Air Blank		0.000	01:12	
Control Test		0.079	01:13	
Air Blank		0.000	01:13	
Diagnostics Check		OK	01:13	

Data Download Date/Time: 05/27/2018 16:03

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROE,ROBERT,D
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SALVADOR,ELDER
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 04/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: CARSON

DATE/TIME OF TEST: 04/29/2018 01:28:06

Diagnostics Check		OK	01:24	
Air Blank		0.000	01:24	
Control Test		0.080	01:24	
Air Blank		0.000	01:25	
Subject Sample	# 1		01:26	Subject Test Refused
Breath Volume		0.000	01:26	
Air Blank		0.000	01:27	
Control Test		0.079	01:27	
Air Blank		0.000	01:28	
Diagnostics Check		OK	01:28	

Data Download Date/Time: 05/27/2018 16:03

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 03/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: DAZULMA

DATE/TIME OF TEST: 04/04/2018 01:31:58

Diagnostics Check		OK	01:23
Air Blank		0.000	01:24
Control Test		0.078	01:24
Air Blank		0.000	01:25
Subject Sample	# 1	0.246	01:27
Breath Volume		1.578	01:27
Air Blank		0.000	01:27
Air Blank		0.000	01:29
Subject Sample	# 2	0.230	01:30
Breath Volume		1.957	01:30
Air Blank		0.000	01:31
Control Test		0.078	01:31
Air Blank		0.000	01:31
Diagnostics Check		OK	01:31

Data Download Date/Time: 04/26/2018 15:33

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 03/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: DANIELS

DATE/TIME OF TEST: 04/07/2018 23:33:08

Diagnostics Check		OK	23:26
Air Blank		0.000	23:27
Control Test		0.079	23:27
Air Blank		0.000	23:28
Subject Sample	# 1	0.111	23:28
Breath Volume		1.644	23:28
Air Blank		0.000	23:29
Air Blank		0.000	23:31
Subject Sample	# 2	0.109	23:31
Breath Volume		1.441	23:31
Air Blank		0.000	23:32
Control Test		0.080	23:32
Air Blank		0.000	23:33
Diagnostics Check		OK	23:33

Data Download Date/Time: 04/26/2018 15:33

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 03/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 04/09/2018 02:13:33

Diagnostics Check		OK	02:04
Air Blank		0.000	02:04
Control Test		0.079	02:05
Air Blank		0.000	02:05
Subject Sample	# 1	0.098	02:08
Breath Volume		1.937	02:08
Air Blank		0.000	02:08
Air Blank		0.000	02:10
Subject Sample	# 2	0.090	02:12
Breath Volume		1.210	02:12
Air Blank		0.000	02:12
Control Test		0.079	02:12
Air Blank		0.000	02:13
Diagnostics Check		OK	02:13

Data Download Date/Time: 04/26/2018 15:33

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR',ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 03/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: RAMBEAU

DATE/TIME OF TEST: 04/10/2018 02:48:56

Diagnostics Check		OK	02:46	
Air Blank		0.000	02:46	
Control Test		0.079	02:46	
Air Blank		0.000	02:47	
Subject Sample	# 1		02:47	Subject Test Refused
Breath Volume		0.000	02:47	
Air Blank		0.000	02:48	
Control Test		0.079	02:48	
Air Blank		0.000	02:48	
Diagnostics Check		OK	02:48	

Data Download Date/Time: 04/26/2018 15:33

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 03/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: WALTERS

DATE/TIME OF TEST: 04/14/2018 00:52:37

Diagnostics Check		OK	00:46
Air Blank		0.000	00:46
Control Test		0.079	00:47
Air Blank		0.000	00:47
Subject Sample	# 1	0.153	00:48
Breath Volume		1.472	00:48
Air Blank		0.000	00:48
Air Blank		0.000	00:50
Subject Sample	# 2	0.148	00:51
Breath Volume		1.617	00:51
Air Blank		0.000	00:51
Control Test		0.080	00:52
Air Blank		0.000	00:52
Diagnostics Check		OK	00:52

Data Download Date/Time: 04/26/2018 15:33

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 03/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 17:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:26

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 04/17/2018 18:01:35

Diagnostics Check		OK	17:58	
Air Blank		0.000	17:59	
Control Test		0.079	17:59	
Air Blank		0.000	18:00	
Subject Sample	# 1		18:00	Subject Test Refused
Breath Volume		0.000	18:00	
Air Blank		0.000	18:00	
Control Test		0.079	18:01	
Air Blank		0.000	18:01	
Diagnostics Check		OK	18:01	

Data Download Date/Time: 04/26/2018 15:33

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 03/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 17:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:26

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 04/17/2018 18:12:39

Diagnostics Check		OK	18:03
Air Blank		0.000	18:03
Control Test		0.080	18:04
Air Blank		0.000	18:04
Subject Sample	# 1	0.176	18:05
Breath Volume		2.843	18:05
Air Blank		0.000	18:05
Air Blank		0.000	18:07
Subject Sample	# 2	0.155	18:08
Breath Volume		1.187	18:08
Air Blank		0.000	18:08
Air Blank		0.000	18:10
Subject Sample	# 3	0.175	18:11
Breath Volume		3.054	18:11
Air Blank		0.000	18:11
Control Test		0.078	18:12
Air Blank		0.000	18:12
Diagnostics Check		OK	18:12

Data Download Date/Time: 04/26/2018 15:33

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 03/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: HARRIOTT

DATE/TIME OF TEST: 04/19/2018 01:42:11

Diagnostics Check		OK	01:36
Air Blank		0.000	01:36
Control Test		0.079	01:37
Air Blank		0.000	01:37
Subject Sample	# 1	0.111	01:37
Breath Volume		1.140	01:37
Air Blank		0.000	01:38
Air Blank		0.000	01:40
Subject Sample	# 2	0.109	01:40
Breath Volume		2.246	01:40
Air Blank		0.000	01:41
Control Test		0.078	01:41
Air Blank		0.000	01:42
Diagnostics Check		OK	01:42

Data Download Date/Time: 04/26/2018 15:33

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 03/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: CARVAJAL

DATE/TIME OF TEST: 04/23/2018 02:16:37

Diagnostics Check		OK	02:13	
Air Blank		0.000	02:14	
Control Test		0.078	02:14	
Air Blank		0.000	02:15	
Subject Sample	# 1		02:15	Subject Test Refused
Breath Volume		0.000	02:15	
Air Blank		0.000	02:15	
Control Test		0.079	02:16	
Air Blank		0.000	02:16	
Diagnostics Check		OK	02:16	

Data Download Date/Time: 04/26/2018 15:33

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 04/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: GIL

DATE/TIME OF TEST: 04/27/2018 02:32:04

Diagnostics Check		OK	02:29	
Air Blank		0.000	02:29	
Control Test		0.080	02:30	
Air Blank		0.000	02:30	
Subject Sample	# 1		02:30	Subject Test Refused
Breath Volume		0.000	02:30	
Air Blank		0.000	02:31	
Control Test		0.080	02:31	
Air Blank		0.000	02:31	
Diagnostics Check		OK	02:32	

Data Download Date/Time: 05/30/2018 10:40

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 04/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 19:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: RABELL

DATE/TIME OF TEST: 04/27/2018 21:01:37

Diagnostics Check		OK	20:58	
Air Blank		0.000	20:58	
Control Test		0.079	20:59	
Air Blank		0.000	20:59	
Subject Sample	# 1		21:00	Subject Test Refused
Breath Volume		0.000	21:00	
Air Blank		0.000	21:00	
Control Test		0.079	21:01	
Air Blank		0.000	21:01	
Diagnostics Check		OK	21:01	

Data Download Date/Time: 05/30/2018 10:40

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEISTHAMMEL,SEAN,P
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: CALDERON,ANDREA
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 04/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: DAVID

DATE/TIME OF TEST: 04/29/2018 23:51:12

Diagnostics Check		OK	23:44
Air Blank		0.000	23:45
Control Test		0.079	23:45
Air Blank		0.000	23:46
Subject Sample	# 1	0.267	23:46
Breath Volume		2.074	23:46
Air Blank		0.000	23:47
Air Blank		0.000	23:49
Subject Sample	# 2	0.251	23:49
Breath Volume		1.214	23:49
Air Blank		0.000	23:50
Control Test		0.078	23:50
Air Blank		0.000	23:51
Diagnostics Check		OK	23:51

Data Download Date/Time: 05/30/2018 10:40

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,GAVIN,O
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: HENRY,GAVIN
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 04/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:23

SUBJECT'S LAST NAME: BARR

DATE/TIME OF TEST: 04/30/2018 22:58:05

Diagnostics Check		OK	22:55	
Air Blank		0.000	22:55	
Control Test		0.079	22:56	
Air Blank		0.000	22:56	
Subject Sample	# 1		22:56	RFI Detect
Breath Volume		0.000	22:56	
Air Blank		0.000	22:57	
Control Test		0.080	22:57	
Air Blank		0.000	22:58	
Diagnostics Check		OK	22:58	

Data Download Date/Time: 05/30/2018 10:40

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACK,C,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BROWN,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: BARELA

DATE/TIME OF TEST: 04/05/2018 21:53:29

Diagnostics Check		OK	21:50	
Air Blank		0.000	21:51	
Control Test		0.079	21:51	
Air Blank		0.000	21:51	
Subject Sample	# 1		21:52	Subject Test Refused
Breath Volume		0.000	21:52	
Air Blank		0.000	21:52	
Control Test		0.079	21:52	
Air Blank		0.000	21:53	
Diagnostics Check		OK	21:53	

Data Download Date/Time: 04/17/2018 12:52

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACK,C,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WEBER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: BURTON

DATE/TIME OF TEST: 04/05/2018 22:29:14

Diagnostics Check		OK	22:26	
Air Blank		0.000	22:26	
Control Test		0.080	22:27	
Air Blank		0.000	22:27	
Subject Sample	# 1		22:27	Subject Test Refused
Breath Volume		0.000	22:27	
Air Blank		0.000	22:28	
Control Test		0.079	22:28	
Air Blank		0.000	22:29	
Diagnostics Check		OK	22:29	

Data Download Date/Time: 04/17/2018 12:52

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BROWN,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: DOROTEO SANTIAGO

DATE/TIME OF TEST: 04/09/2018 00:47:38

Diagnostics Check		OK	00:41
Air Blank		0.000	00:41
Control Test		0.080	00:42
Air Blank		0.000	00:42
Subject Sample	# 1	0.087	00:43
Breath Volume		1.781	00:43
Air Blank		0.000	00:43
Air Blank		0.000	00:45
Subject Sample	# 2	0.088	00:46
Breath Volume		2.808	00:46
Air Blank		0.000	00:46
Control Test		0.079	00:47
Air Blank		0.000	00:47
Diagnostics Check		OK	00:47

Data Download Date/Time: 04/17/2018 12:52

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRYANT,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: ESTES

DATE/TIME OF TEST: 04/09/2018 02:17:16

Diagnostics Check		OK	02:13	
Air Blank		0.000	02:14	
Control Test		0.080	02:14	
Air Blank		0.000	02:15	
Subject Sample	# 1		02:15	Subject Test Refused
Breath Volume		0.000	02:15	
Air Blank		0.000	02:16	
Control Test		0.080	02:16	
Air Blank		0.000	02:17	
Diagnostics Check		OK	02:17	

Data Download Date/Time: 04/17/2018 12:52

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HYSMITH,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: VAN MOURIK

DATE/TIME OF TEST: 04/09/2018 04:13:00

Diagnostics Check		OK	04:10	
Air Blank		0.000	04:10	
Control Test		0.080	04:10	
Air Blank		0.000	04:11	
Subject Sample	# 1		04:11	Subject Test Refused
Breath Volume		0.000	04:11	
Air Blank		0.000	04:12	
Control Test		0.080	04:12	
Air Blank		0.000	04:12	
Diagnostics Check		OK	04:13	

Data Download Date/Time: 04/17/2018 12:52

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: RODRIQUEZ

DATE/TIME OF TEST: 04/12/2018 01:49:48

Diagnostics Check		OK	01:43
Air Blank		0.000	01:43
Control Test		0.080	01:44
Air Blank		0.000	01:44
Subject Sample	# 1	0.188	01:45
Breath Volume		3.097	01:45
Air Blank		0.000	01:45
Air Blank		0.000	01:47
Subject Sample	# 2	0.181	01:48
Breath Volume		3.238	01:48
Air Blank		0.000	01:48
Control Test		0.079	01:49
Air Blank		0.000	01:49
Diagnostics Check		OK	01:49

Data Download Date/Time: 04/17/2018 12:52

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACK,C,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BROWN,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: RHODES

DATE/TIME OF TEST: 04/15/2018 21:43:56

Diagnostics Check		OK	21:41	
Air Blank		0.000	21:41	
Control Test		0.080	21:41	
Air Blank		0.000	21:42	
Subject Sample	# 1		21:42	Subject Test Refused
Breath Volume		0.000	21:42	
Air Blank		0.000	21:43	
Control Test		0.079	21:43	
Air Blank		0.000	21:43	
Diagnostics Check		OK	21:43	

Data Download Date/Time: 04/17/2018 12:52

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACK,C,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRYANT,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 04/15/2018 23:02:59

Diagnostics Check		OK	22:55
Air Blank		0.000	22:56
Control Test		0.079	22:56
Air Blank		0.000	22:57
Subject Sample	# 1	0.222	22:57
Breath Volume		1.445	22:57
Air Blank		0.000	22:58
Air Blank		0.000	23:00
Subject Sample	# 2	0.209	23:01
Breath Volume		2.023	23:01
Air Blank		0.000	23:02
Control Test		0.078	23:02
Air Blank		0.000	23:02
Diagnostics Check		OK	23:02

Data Download Date/Time: 04/17/2018 12:53

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: NEESON

DATE/TIME OF TEST: 04/22/2018 03:39:38

Diagnostics Check		OK	03:33
Air Blank		0.000	03:33
Control Test		0.080	03:34
Air Blank		0.000	03:34
Subject Sample	# 1	0.043	03:35
Breath Volume		2.691	03:35
Air Blank		0.000	03:35
Air Blank		0.000	03:37
Subject Sample	# 2	0.045	03:38
Breath Volume		2.578	03:38
Air Blank		0.000	03:38
Control Test		0.079	03:39
Air Blank		0.000	03:39
Diagnostics Check		OK	03:39

Data Download Date/Time: 05/01/2018 14:54

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WEBER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 04/23/2018 02:08:56

Diagnostics Check		OK	02:06	
Air Blank		0.000	02:06	
Control Test		0.080	02:06	
Air Blank		0.000	02:07	
Subject Sample	# 1		02:07	Subject Test Refused
Breath Volume		0.000	02:07	
Air Blank		0.000	02:08	
Control Test		0.079	02:08	
Air Blank		0.000	02:08	
Diagnostics Check		OK	02:08	

Data Download Date/Time: 05/01/2018 14:54

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RUSSELL,C
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 04/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: ORTEGA MARTINEZ

DATE/TIME OF TEST: 04/23/2018 04:42:59

Diagnostics Check		OK	04:35
Air Blank		0.000	04:35
Control Test		0.079	04:36
Air Blank		0.000	04:36
Subject Sample	# 1	0.149	04:37
Breath Volume		1.339	04:37
Air Blank		0.000	04:38
Air Blank		0.000	04:40
Subject Sample	# 2	0.146	04:41
Breath Volume		1.402	04:41
Air Blank		0.000	04:42
Control Test		0.078	04:42
Air Blank		0.000	04:42
Diagnostics Check		OK	04:42

Data Download Date/Time: 05/01/2018 14:54

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ZURN,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: REECE

DATE/TIME OF TEST: 04/24/2018 01:15:57

Diagnostics Check		OK	01:09
Air Blank		0.000	01:10
Control Test		0.080	01:10
Air Blank		0.000	01:11
Subject Sample	# 1	0.081	01:11
Breath Volume		1.234	01:11
Air Blank		0.000	01:12
Air Blank		0.000	01:14
Subject Sample	# 2	0.083	01:14
Breath Volume		1.125	01:14
Air Blank		0.000	01:15
Control Test		0.079	01:15
Air Blank		0.000	01:15
Diagnostics Check		OK	01:15

Data Download Date/Time: 05/01/2018 14:54

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DIAZ,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: COCANOUGHER

DATE/TIME OF TEST: 04/24/2018 04:51:49

Diagnostics Check		OK	04:48	
Air Blank		0.000	04:49	
Control Test		0.080	04:49	
Air Blank		0.000	04:50	
Subject Sample	# 1		04:50	Subject Test Refused
Breath Volume		0.000	04:50	
Air Blank		0.000	04:50	
Control Test		0.080	04:51	
Air Blank		0.000	04:51	
Diagnostics Check		OK	04:51	

Data Download Date/Time: 05/01/2018 14:54

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACK,C,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BROWN,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:34

SUBJECT'S LAST NAME: FITZPATRICK

DATE/TIME OF TEST: 04/25/2018 19:01:03

Diagnostics Check		OK	18:54
Air Blank		0.000	18:54
Control Test		0.079	18:55
Air Blank		0.000	18:55
Subject Sample	# 1	0.218	18:56
Breath Volume		1.363	18:56
Air Blank		0.000	18:56
Air Blank		0.000	18:58
Subject Sample	# 2	0.211	18:59
Breath Volume		1.437	18:59
Air Blank		0.000	19:00
Control Test		0.079	19:00
Air Blank		0.000	19:00
Diagnostics Check		OK	19:01

Data Download Date/Time: 05/01/2018 14:54

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACK,C,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BROWN,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 04/26/2018 22:15:22

Diagnostics Check		OK	22:12	
Air Blank		0.000	22:12	
Control Test		0.079	22:13	
Air Blank		0.000	22:13	
Subject Sample	# 1		22:13	Subject Test Refused
Breath Volume		0.000	22:13	
Air Blank		0.000	22:14	
Control Test		0.079	22:14	
Air Blank		0.000	22:15	
Diagnostics Check		OK	22:15	

Data Download Date/Time: 05/01/2018 14:54

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNOW,CHRIS,D
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: THEUS,J
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 03/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: LANE

DATE/TIME OF TEST: 04/14/2018 02:31:19

Diagnostics Check		OK	02:27	
Air Blank		0.000	02:28	
Control Test		0.081	02:28	
Air Blank		0.000	02:29	
Subject Sample	# 1	0.183	02:29	
Breath Volume		1.519	02:29	
Air Blank			02:29	RFI Detect
Air Blank		0.000	02:30	
Control Test		0.080	02:30	
Air Blank		0.000	02:31	
Diagnostics Check		OK	02:31	

Data Download Date/Time: 04/21/2018 14:49

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNOW,CHRIS,D
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JETT,C
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: DEARBORNE

DATE/TIME OF TEST: 04/14/2018 04:39:50

Diagnostics Check		OK	04:33
Air Blank		0.000	04:33
Control Test		0.081	04:34
Air Blank		0.000	04:34
Subject Sample	# 1	0.091	04:35
Breath Volume		1.121	04:35
Air Blank		0.000	04:35
Air Blank		0.000	04:37
Subject Sample	# 2	0.091	04:38
Breath Volume		1.453	04:38
Air Blank		0.000	04:38
Control Test		0.080	04:39
Air Blank		0.000	04:39
Diagnostics Check		OK	04:39

Data Download Date/Time: 04/21/2018 14:49

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNOW,C,D
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: PHILLIPS,Z
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: BORYS

DATE/TIME OF TEST: 04/14/2018 05:25:43

Diagnostics Check		OK	05:22	
Air Blank		0.000	05:23	
Control Test		0.081	05:23	
Air Blank		0.000	05:24	
Subject Sample	# 1		05:24	Subject Test Refused
Breath Volume		0.000	05:24	
Air Blank		0.000	05:24	
Control Test		0.081	05:25	
Air Blank		0.000	05:25	
Diagnostics Check		OK	05:25	

Data Download Date/Time: 04/21/2018 14:49

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNOW,CHRIS,D
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: HYSLER,D
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: CARR

DATE/TIME OF TEST: 04/14/2018 20:34:16

Diagnostics Check	OK	20:32	
Air Blank	0.000	20:33	
Control Test	0.080	20:33	
Air Blank		20:33	RFI Detect
Air Blank		20:34	RFI Detect

Data Download Date/Time: 04/21/2018 14:49

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNOW,CHRIS,D
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: HYSLER,D
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: CARR

DATE/TIME OF TEST: 04/14/2018 20:47:17

Diagnostics Check		OK	20:35	
Air Blank		0.000	20:36	
Control Test		0.081	20:36	
Air Blank		0.000	20:37	
Subject Sample	# 1	0.123	20:40	Volume Not Met
Breath Volume		0.406	20:40	
Air Blank		0.000	20:40	
Air Blank		0.000	20:42	
Subject Sample	# 2	0.107	20:45	Volume Not Met
Breath Volume		0.539	20:45	
Air Blank		0.000	20:46	
Control Test		0.079	20:46	
Air Blank		0.000	20:47	
Diagnostics Check		OK	20:47	

Data Download Date/Time: 04/21/2018 14:49

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAUNDERS,DEPUTY,C
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: COX,DEPUTY
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 15:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:20

SUBJECT'S LAST NAME: HAISTEN

DATE/TIME OF TEST: 04/15/2018 16:48:53

Diagnostics Check		OK	16:41
Air Blank		0.000	16:42
Control Test		0.080	16:42
Air Blank		0.000	16:43
Subject Sample	# 1	0.075	16:43
Breath Volume		1.464	16:43
Air Blank		0.000	16:44
Air Blank		0.000	16:46
Subject Sample	# 2	0.072	16:47
Breath Volume		1.578	16:47
Air Blank		0.000	16:48
Control Test		0.080	16:48
Air Blank		0.000	16:48
Diagnostics Check		OK	16:48

Data Download Date/Time: 04/21/2018 14:49

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAUNLICH,G,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: FREHULFER,S
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 04/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 04/22/2018 03:27:47

Diagnostics Check		OK	03:16	
Air Blank		0.000	03:16	
Control Test		0.080	03:17	
Air Blank		0.000	03:17	
Subject Sample	# 1		03:20	No Sample Provided
Breath Volume		0.000	03:20	
Air Blank		0.000	03:21	
Air Blank		0.000	03:23	
Subject Sample	# 2		03:26	No Sample Provided
Breath Volume		0.000	03:26	
Air Blank		0.000	03:26	
Control Test		0.081	03:27	
Air Blank		0.000	03:27	
Diagnostics Check		OK	03:27	

Data Download Date/Time: 05/19/2018 14:52

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,DAKOTA,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: STIVERS,H
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: FLOREZ

DATE/TIME OF TEST: 04/23/2018 01:25:53

Diagnostics Check		OK	01:19
Air Blank		0.000	01:20
Control Test		0.080	01:20
Air Blank		0.000	01:21
Subject Sample	# 1	0.184	01:21
Breath Volume		1.664	01:21
Air Blank		0.000	01:21
Air Blank		0.000	01:23
Subject Sample	# 2	0.182	01:24
Breath Volume		1.875	01:24
Air Blank		0.000	01:24
Control Test		0.080	01:25
Air Blank		0.000	01:25
Diagnostics Check		OK	01:25

Data Download Date/Time: 05/19/2018 14:52

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAUNLICH,G,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: HESTER

DATE/TIME OF TEST: 04/27/2018 01:44:50

Diagnostics Check		OK	01:37
Air Blank		0.000	01:37
Control Test		0.081	01:38
Air Blank		0.000	01:38
Subject Sample	# 1	0.281	01:39
Breath Volume		1.863	01:39
Air Blank		0.000	01:40
Air Blank		0.000	01:42
Subject Sample	# 2	0.282	01:43
Breath Volume		1.832	01:43
Air Blank		0.000	01:43
Control Test		0.080	01:44
Air Blank		0.000	01:44
Diagnostics Check		OK	01:44

Data Download Date/Time: 05/19/2018 14:52

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FITHIAN,JASON
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: BIRGS

DATE/TIME OF TEST: 04/01/2018 02:45:12

Diagnostics Check		OK	02:39
Air Blank		0.000	02:39
Control Test		0.079	02:39
Air Blank		0.000	02:40
Subject Sample	# 1	0.034	02:40
Breath Volume		1.453	02:40
Air Blank		0.000	02:41
Air Blank		0.000	02:43
Subject Sample	# 2	0.032	02:43
Breath Volume		1.464	02:43
Air Blank		0.000	02:44
Control Test		0.080	02:44
Air Blank		0.000	02:45
Diagnostics Check		OK	02:45

Data Download Date/Time: 04/26/2018 12:11

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 04/01/2018 05:08:31

Diagnostics Check		OK	05:02
Air Blank		0.000	05:02
Control Test		0.080	05:02
Air Blank		0.000	05:03
Subject Sample	# 1	0.138	05:03
Breath Volume		2.218	05:03
Air Blank		0.000	05:04
Air Blank		0.000	05:06
Subject Sample	# 2	0.127	05:07
Breath Volume		1.957	05:07
Air Blank		0.000	05:07
Control Test		0.080	05:07
Air Blank		0.000	05:08
Diagnostics Check		OK	05:08

Data Download Date/Time: 04/26/2018 12:11

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: REYNOLDS,JOSHUA
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: SCRUGGS

DATE/TIME OF TEST: 04/02/2018 05:16:21

Diagnostics Check		OK	05:09
Air Blank		0.000	05:10
Control Test		0.080	05:10
Air Blank		0.000	05:11
Subject Sample	# 1	0.140	05:11
Breath Volume		3.140	05:11
Air Blank		0.000	05:12
Air Blank		0.000	05:14
Subject Sample	# 2	0.134	05:14
Breath Volume		2.332	05:14
Air Blank		0.000	05:15
Control Test		0.080	05:15
Air Blank		0.000	05:16
Diagnostics Check		OK	05:16

Data Download Date/Time: 04/26/2018 12:11

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: TRABUE,THOMAS
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: FARRAKHAN

DATE/TIME OF TEST: 04/05/2018 04:34:17

Diagnostics Check		OK	04:27
Air Blank		0.000	04:27
Control Test		0.081	04:28
Air Blank		0.000	04:28
Subject Sample	# 1	0.162	04:29
Breath Volume		2.617	04:29
Air Blank		0.000	04:30
Air Blank		0.000	04:32
Subject Sample	# 2	0.159	04:32
Breath Volume		2.558	04:32
Air Blank		0.000	04:33
Control Test		0.081	04:33
Air Blank		0.000	04:34
Diagnostics Check		OK	04:34

Data Download Date/Time: 04/26/2018 12:11

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: TAVARES DE OLIVEIRA

DATE/TIME OF TEST: 04/06/2018 03:19:36

Diagnostics Check		OK	03:13
Air Blank		0.000	03:13
Control Test		0.081	03:14
Air Blank		0.000	03:14
Subject Sample	# 1	0.225	03:15
Breath Volume		1.953	03:15
Air Blank		0.000	03:15
Air Blank		0.000	03:17
Subject Sample	# 2	0.227	03:18
Breath Volume		1.355	03:18
Air Blank		0.000	03:18
Control Test		0.080	03:19
Air Blank		0.000	03:19
Diagnostics Check		OK	03:19

Data Download Date/Time: 04/26/2018 12:11

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KEATING,KAYLA
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: BRANDT

DATE/TIME OF TEST: 04/07/2018 04:40:39

Diagnostics Check		OK	04:29	
Air Blank		0.000	04:29	
Control Test		0.081	04:30	
Air Blank		0.000	04:30	
Subject Sample	# 1		04:33	No Sample Provided
Breath Volume		0.000	04:33	
Air Blank		0.000	04:34	
Air Blank		0.000	04:36	
Subject Sample	# 2		04:39	No Sample Provided
Breath Volume		0.000	04:39	
Air Blank		0.000	04:39	
Control Test		0.080	04:40	
Air Blank		0.000	04:40	
Diagnostics Check		OK	04:40	

Data Download Date/Time: 04/26/2018 12:11

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KEATING,KAYLA
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 04/08/2018 03:57:41

Diagnostics Check		OK	03:51
Air Blank		0.000	03:51
Control Test		0.081	03:52
Air Blank		0.000	03:52
Subject Sample	# 1	0.285	03:53
Breath Volume		1.839	03:53
Air Blank		0.000	03:53
Air Blank		0.000	03:55
Subject Sample	# 2	0.287	03:56
Breath Volume		1.589	03:56
Air Blank		0.000	03:56
Control Test		0.080	03:57
Air Blank		0.000	03:57
Diagnostics Check		OK	03:57

Data Download Date/Time: 04/26/2018 12:11

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HENNESSY,JONATHAN
ARRESTING OFFICER AGENCY: Sunrise PD
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: JARAMILLO

DATE/TIME OF TEST: 04/08/2018 04:27:55

Diagnostics Check		OK	04:19
Air Blank		0.000	04:19
Control Test		0.081	04:20
Air Blank		0.000	04:20
Subject Sample	# 1	0.202	04:21
Breath Volume		1.636	04:21
Air Blank		0.000	04:22
Air Blank		0.000	04:24
Subject Sample	# 2	0.213	04:26
Breath Volume		1.796	04:26
Air Blank		0.000	04:27
Control Test		0.080	04:27
Air Blank		0.000	04:27
Diagnostics Check		OK	04:27

Data Download Date/Time: 04/26/2018 12:11

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: AMENGOR,JESSICA
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 05:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:39

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 04/08/2018 06:06:17

Diagnostics Check		OK	05:59
Air Blank		0.000	06:00
Control Test		0.080	06:00
Air Blank		0.000	06:00
Subject Sample	# 1	0.185	06:01
Breath Volume		1.210	06:01
Air Blank		0.000	06:02
Air Blank		0.000	06:04
Subject Sample	# 2	0.188	06:04
Breath Volume		1.164	06:04
Air Blank		0.000	06:05
Control Test		0.081	06:05
Air Blank		0.000	06:06
Diagnostics Check		OK	06:06

Data Download Date/Time: 04/26/2018 12:12

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MARINO,GREGORY
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: LIGHTFOOT

DATE/TIME OF TEST: 04/09/2018 00:41:46

Diagnostics Check		OK	00:33
Air Blank		0.000	00:33
Control Test		0.080	00:33
Air Blank		0.000	00:34
Subject Sample	# 1	0.219	00:36
Breath Volume		2.191	00:36
Air Blank		0.000	00:37
Air Blank		0.000	00:39
Subject Sample	# 2	0.225	00:40
Breath Volume		1.925	00:40
Air Blank		0.000	00:40
Control Test		0.080	00:41
Air Blank		0.000	00:41
Diagnostics Check		OK	00:41

Data Download Date/Time: 04/26/2018 12:12

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LOUGHRAN,KEVIN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 04:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: KIM

DATE/TIME OF TEST: 04/11/2018 05:48:41

Diagnostics Check		OK	05:40
Air Blank		0.000	05:41
Control Test		0.082	05:41
Air Blank		0.000	05:42
Subject Sample	# 1	0.186	05:44
Breath Volume		1.242	05:44
Air Blank		0.000	05:44
Air Blank		0.000	05:46
Subject Sample	# 2	0.187	05:47
Breath Volume		1.707	05:47
Air Blank		0.000	05:47
Control Test		0.079	05:48
Air Blank		0.000	05:48
Diagnostics Check		OK	05:48

Data Download Date/Time: 04/26/2018 12:12

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SACKMANN,ANDREW
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 15:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: DALGLEISH

DATE/TIME OF TEST: 04/12/2018 17:41:07

Diagnostics Check		OK	17:35
Air Blank		0.000	17:35
Control Test		0.081	17:35
Air Blank		0.000	17:36
Subject Sample	# 1	0.000	17:36
Breath Volume		1.425	17:36
Air Blank		0.000	17:37
Air Blank		0.000	17:39
Subject Sample	# 2	0.000	17:39
Breath Volume		1.515	17:39
Air Blank		0.000	17:40
Control Test		0.081	17:40
Air Blank		0.000	17:41
Diagnostics Check		OK	17:41

Data Download Date/Time: 04/26/2018 12:12

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: D'ONOFRIO

DATE/TIME OF TEST: 04/14/2018 02:56:30

Diagnostics Check		OK	02:48
Air Blank		0.000	02:48
Control Test		0.080	02:49
Air Blank		0.000	02:49
Subject Sample	# 1	0.159	02:51
Breath Volume		3.312	02:51
Air Blank		0.000	02:52
Air Blank		0.000	02:54
Subject Sample	# 2	0.154	02:55
Breath Volume		2.453	02:55
Air Blank		0.000	02:55
Control Test		0.079	02:55
Air Blank		0.000	02:56
Diagnostics Check		OK	02:56

Data Download Date/Time: 04/26/2018 12:12

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: PACK,JEFFREY
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 05:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:03

SUBJECT'S LAST NAME: TEJADA

DATE/TIME OF TEST: 04/14/2018 06:38:13

Diagnostics Check		OK	06:30
Air Blank		0.000	06:31
Control Test		0.080	06:31
Air Blank		0.000	06:32
Subject Sample	# 1	0.157	06:33
Breath Volume		1.804	06:33
Air Blank		0.000	06:34
Air Blank		0.000	06:36
Subject Sample	# 2	0.152	06:36
Breath Volume		1.335	06:36
Air Blank		0.000	06:37
Control Test		0.080	06:37
Air Blank		0.000	06:38
Diagnostics Check		OK	06:38

Data Download Date/Time: 04/26/2018 12:12

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: STRZALKOWSKI,SEAN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: AGUASVIVAS

DATE/TIME OF TEST: 04/18/2018 02:30:15

Diagnostics Check		OK	02:23
Air Blank		0.000	02:23
Control Test		0.081	02:24
Air Blank		0.000	02:24
Subject Sample	# 1	0.149	02:25
Breath Volume		1.179	02:25
Air Blank		0.000	02:25
Air Blank		0.000	02:27
Subject Sample	# 2	0.143	02:28
Breath Volume		2.281	02:28
Air Blank		0.000	02:29
Control Test		0.080	02:29
Air Blank		0.000	02:30
Diagnostics Check		OK	02:30

Data Download Date/Time: 04/26/2018 12:12

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 04:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:42

SUBJECT'S LAST NAME: KOSAR

DATE/TIME OF TEST: 04/19/2018 06:10:54

Diagnostics Check		OK	06:02
Air Blank		0.000	06:02
Control Test		0.081	06:03
Air Blank		0.000	06:03
Subject Sample	# 1	0.108	06:06
Breath Volume		1.406	06:06
Air Blank		0.000	06:06
Air Blank		0.000	06:08
Subject Sample	# 2	0.105	06:09
Breath Volume		1.222	06:09
Air Blank		0.000	06:10
Control Test		0.080	06:10
Air Blank		0.000	06:10
Diagnostics Check		OK	06:10

Data Download Date/Time: 04/26/2018 12:12

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KAPLAN,SAMANTHA
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: ALEXANDER

DATE/TIME OF TEST: 04/20/2018 00:34:08

Diagnostics Check		OK	00:28
Air Blank		0.000	00:28
Control Test		0.081	00:28
Air Blank		0.000	00:29
Subject Sample	# 1	0.124	00:29
Breath Volume		2.640	00:29
Air Blank		0.000	00:30
Air Blank		0.000	00:32
Subject Sample	# 2	0.116	00:32
Breath Volume		2.878	00:32
Air Blank		0.000	00:33
Control Test		0.080	00:33
Air Blank		0.000	00:34
Diagnostics Check		OK	00:34

Data Download Date/Time: 04/26/2018 12:12

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: HORNSTEIN

DATE/TIME OF TEST: 04/20/2018 04:14:06

Diagnostics Check		OK	04:07
Air Blank		0.000	04:08
Control Test		0.081	04:08
Air Blank		0.000	04:09
Subject Sample	# 1	0.062	04:09
Breath Volume		2.125	04:09
Air Blank		0.000	04:10
Air Blank		0.000	04:12
Subject Sample	# 2	0.056	04:12
Breath Volume		1.203	04:12
Air Blank		0.000	04:13
Control Test		0.080	04:13
Air Blank		0.000	04:14
Diagnostics Check		OK	04:14

Data Download Date/Time: 04/26/2018 12:12

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BOHM,JOHN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: SILVA

DATE/TIME OF TEST: 04/20/2018 05:11:12

Diagnostics Check		OK	04:59	
Air Blank		0.000	05:00	
Control Test		0.080	05:00	
Air Blank		0.000	05:01	
Subject Sample	# 1		05:04	No Sample Provided
Breath Volume		0.000	05:04	
Air Blank		0.000	05:04	
Air Blank		0.000	05:06	
Subject Sample	# 2	0.150	05:09	Volume Not Met
Breath Volume		0.523	05:09	
Air Blank		0.000	05:10	
Control Test		0.079	05:10	
Air Blank		0.000	05:11	
Diagnostics Check		OK	05:11	

Data Download Date/Time: 04/26/2018 12:12

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BOHM,JOHN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: SILVA

DATE/TIME OF TEST: 04/20/2018 05:24:03

Diagnostics Check		OK	05:15
Air Blank		0.000	05:15
Control Test		0.079	05:15
Air Blank		0.000	05:16
Subject Sample	# 1	0.159	05:19
Breath Volume		1.312	05:19
Air Blank		0.000	05:20
Air Blank		0.000	05:22
Subject Sample	# 2	0.165	05:22
Breath Volume		2.050	05:22
Air Blank		0.000	05:23
Control Test		0.079	05:23
Air Blank		0.000	05:23
Diagnostics Check		OK	05:24

Data Download Date/Time: 04/26/2018 12:12

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BRENTON,JESSIKA
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 04:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:07

SUBJECT'S LAST NAME: GUERRERO ANDINO

DATE/TIME OF TEST: 04/20/2018 06:34:50

Diagnostics Check		OK	06:27
Air Blank		0.000	06:28
Control Test		0.080	06:28
Air Blank		0.000	06:29
Subject Sample	# 1	0.245	06:30
Breath Volume		2.265	06:30
Air Blank		0.000	06:30
Air Blank		0.000	06:32
Subject Sample	# 2	0.244	06:33
Breath Volume		2.421	06:33
Air Blank		0.000	06:33
Control Test		0.080	06:34
Air Blank		0.000	06:34
Diagnostics Check		OK	06:34

Data Download Date/Time: 04/26/2018 12:12

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ROJAS,RANDY
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: GUARDIA

DATE/TIME OF TEST: 04/21/2018 01:37:23

Diagnostics Check		OK	01:30
Air Blank		0.000	01:31
Control Test		0.081	01:31
Air Blank		0.000	01:31
Subject Sample	# 1	0.131	01:33
Breath Volume		1.449	01:33
Air Blank		0.000	01:33
Air Blank		0.000	01:35
Subject Sample	# 2	0.138	01:35
Breath Volume		1.437	01:35
Air Blank		0.000	01:36
Control Test		0.080	01:36
Air Blank		0.000	01:37
Diagnostics Check		OK	01:37

Data Download Date/Time: 04/26/2018 12:12

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HADLEY,KIMBERLY
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 07:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:35

SUBJECT'S LAST NAME: SOSA

DATE/TIME OF TEST: 04/22/2018 09:47:14

Diagnostics Check		OK	09:40
Air Blank		0.000	09:41
Control Test		0.081	09:41
Air Blank		0.000	09:41
Subject Sample	# 1	0.170	09:42
Breath Volume		2.609	09:42
Air Blank		0.000	09:43
Air Blank		0.000	09:45
Subject Sample	# 2	0.166	09:45
Breath Volume		2.710	09:45
Air Blank		0.000	09:46
Control Test		0.081	09:46
Air Blank		0.000	09:47
Diagnostics Check		OK	09:47

Data Download Date/Time: 04/26/2018 12:12

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BOHM,JOHN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: PAGOADA HERNANDEZ

DATE/TIME OF TEST: 04/26/2018 02:22:42

Diagnostics Check		OK	02:16
Air Blank		0.000	02:17
Control Test		0.082	02:17
Air Blank		0.000	02:17
Subject Sample	# 1	0.000	02:18
Breath Volume		1.976	02:18
Air Blank		0.000	02:18
Air Blank		0.000	02:20
Subject Sample	# 2	0.000	02:21
Breath Volume		2.484	02:21
Air Blank		0.000	02:21
Control Test		0.081	02:22
Air Blank		0.000	02:22
Diagnostics Check		OK	02:22

Data Download Date/Time: 04/26/2018 12:12

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: TRABUE,THOMAS
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: CARDENA

DATE/TIME OF TEST: 04/26/2018 02:35:49

Diagnostics Check		OK	02:26
Air Blank		0.000	02:27
Control Test		0.080	02:27
Air Blank		0.000	02:27
Subject Sample	# 1	0.173	02:28
Breath Volume		1.792	02:28
Air Blank		0.000	02:29
Air Blank		0.000	02:30
Subject Sample	# 2	0.149	02:31
Breath Volume		1.222	02:31
Air Blank		0.000	02:32
Air Blank		0.000	02:33
Subject Sample	# 3	0.161	02:34
Breath Volume		1.519	02:34
Air Blank		0.000	02:34
Control Test		0.080	02:35
Air Blank		0.000	02:35
Diagnostics Check		OK	02:35

Data Download Date/Time: 04/26/2018 12:12

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURDOCK,TROY,J
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: MURDOCK,TROY
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 03/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: VINDELL

DATE/TIME OF TEST: 04/01/2018 03:36:36

Diagnostics Check		OK	03:30
Air Blank		0.000	03:30
Control Test		0.080	03:30
Air Blank		0.000	03:31
Subject Sample	# 1	0.143	03:32
Breath Volume		4.070	03:32
Air Blank		0.000	03:32
Air Blank		0.000	03:34
Subject Sample	# 2	0.141	03:35
Breath Volume		3.328	03:35
Air Blank		0.000	03:35
Control Test		0.080	03:36
Air Blank		0.000	03:36
Diagnostics Check		OK	03:36

Data Download Date/Time: 04/26/2018 13:40

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAYE,MAKINA,A
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: WAYE,MAKINA
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 03/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 11:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: CHEVERE

DATE/TIME OF TEST: 04/07/2018 00:39:57

Diagnostics Check	OK	00:38	
Air Blank	0.000	00:38	
Control Test	0.079	00:39	
Air Blank	0.000	00:39	
		00:39	RFI Detect
Air Blank		00:39	RFI Detect

Data Download Date/Time: 04/26/2018 13:40

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAYE,MAKINA,A
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: WAYE,MAKINA
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 03/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 11:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: CHEVERE

DATE/TIME OF TEST: 04/07/2018 00:49:05

Diagnostics Check		OK	00:42
Air Blank		0.000	00:43
Control Test		0.079	00:43
Air Blank		0.000	00:44
Subject Sample	# 1	0.127	00:44
Breath Volume		1.605	00:44
Air Blank		0.000	00:45
Air Blank		0.000	00:47
Subject Sample	# 2	0.125	00:47
Breath Volume		2.726	00:47
Air Blank		0.000	00:48
Control Test		0.079	00:48
Air Blank		0.000	00:48
Diagnostics Check		OK	00:49

Data Download Date/Time: 04/26/2018 13:40

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURDOCK,TROY,J
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: PEZZUTO,GIANNI
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 03/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: DELANEY

DATE/TIME OF TEST: 04/09/2018 20:41:46

Diagnostics Check		OK	20:38	
Air Blank		0.000	20:39	
Control Test		0.079	20:39	
Air Blank		0.000	20:40	
Subject Sample	# 1		20:40	Subject Test Refused
Breath Volume		0.000	20:40	
Air Blank		0.000	20:40	
Control Test		0.079	20:41	
Air Blank		0.000	20:41	
Diagnostics Check		OK	20:41	

Data Download Date/Time: 04/26/2018 13:40

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WETHERINGTON,STEPHEN,L
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: GABINO,CARLOS
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 03/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 06:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:16

SUBJECT'S LAST NAME: SALOMONE

DATE/TIME OF TEST: 04/13/2018 07:50:04

Diagnostics Check		OK	07:43
Air Blank		0.000	07:44
Control Test		0.080	07:44
Air Blank		0.000	07:44
Subject Sample	# 1	0.208	07:45
Breath Volume		1.644	07:45
Air Blank		0.000	07:46
Air Blank		0.000	07:47
Subject Sample	# 2	0.215	07:48
Breath Volume		2.367	07:48
Air Blank		0.000	07:49
Control Test		0.079	07:49
Air Blank		0.000	07:49
Diagnostics Check		OK	07:50

Data Download Date/Time: 04/26/2018 13:40

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURDOCK,TROY,J
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: MURDOCK,TROY
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 03/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: ULYANITSKAYA

DATE/TIME OF TEST: 04/14/2018 02:23:54

Diagnostics Check		OK	02:12	
Air Blank		0.000	02:12	
Control Test		0.080	02:13	
Air Blank		0.000	02:13	
Subject Sample	# 1	0.205	02:16	Volume Not Met
Breath Volume		0.031	02:16	
Air Blank		0.000	02:17	
Air Blank		0.000	02:19	
Subject Sample	# 2	0.216	02:22	Volume Not Met
Breath Volume		0.652	02:22	
Air Blank		0.000	02:22	
Control Test		0.079	02:23	
Air Blank		0.000	02:23	
Diagnostics Check		OK	02:23	

Data Download Date/Time: 04/26/2018 13:40

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURDOCK,TROY,J
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: MURDOCK,TROY
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 03/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: ULYANITSKAYA

DATE/TIME OF TEST: 04/14/2018 02:41:48

Diagnostics Check		OK	02:26	
Air Blank		0.000	02:26	
Control Test		0.079	02:26	
Air Blank		0.000	02:27	
Subject Sample	# 1	0.204	02:30	Volume Not Met
Breath Volume		0.230	02:30	
Air Blank		0.000	02:31	
Air Blank		0.000	02:33	
Subject Sample	# 2	0.214	02:35	
Breath Volume		1.210	02:35	
Air Blank		0.000	02:36	
Air Blank		0.000	02:37	
Subject Sample	# 3		02:40	Subject Test Refused
Breath Volume		0.410	02:40	
Air Blank		0.000	02:40	
Control Test		0.078	02:41	
Air Blank		0.000	02:41	
Diagnostics Check		OK	02:41	

Data Download Date/Time: 04/26/2018 13:40

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KABAN,SEAN,W
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: KABAN,SEAN
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 03/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 04/14/2018 03:13:45

Diagnostics Check		OK	03:07
Air Blank		0.000	03:07
Control Test		0.079	03:08
Air Blank		0.000	03:08
Subject Sample	# 1	0.123	03:09
Breath Volume		2.660	03:09
Air Blank		0.000	03:09
Air Blank		0.000	03:11
Subject Sample	# 2	0.127	03:12
Breath Volume		2.039	03:12
Air Blank		0.000	03:12
Control Test		0.080	03:13
Air Blank		0.000	03:13
Diagnostics Check		OK	03:13

Data Download Date/Time: 04/26/2018 13:40

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURDOCK,TROY,J
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: MURDOCK,TROY
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 03/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: PASQUETTE

DATE/TIME OF TEST: 04/15/2018 01:25:43

Diagnostics Check		OK	01:19
Air Blank		0.000	01:19
Control Test		0.079	01:20
Air Blank		0.000	01:20
Subject Sample	# 1	0.108	01:21
Breath Volume		3.464	01:21
Air Blank		0.000	01:21
Air Blank		0.000	01:23
Subject Sample	# 2	0.104	01:24
Breath Volume		2.742	01:24
Air Blank		0.000	01:24
Control Test		0.079	01:25
Air Blank		0.000	01:25
Diagnostics Check		OK	01:25

Data Download Date/Time: 04/26/2018 13:40

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURDOCK,TROY,J
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: KABAN,SEAN
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 03/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: CARRIERO

DATE/TIME OF TEST: 04/15/2018 06:15:03

Diagnostics Check		OK	06:11	
Air Blank		0.000	06:11	
Control Test		0.079	06:11	
Air Blank		0.000	06:12	
Subject Sample	# 1		06:13	Subject Test Refused
Breath Volume		0.000	06:13	
Air Blank		0.000	06:14	
Control Test		0.079	06:14	
Air Blank		0.000	06:14	
Diagnostics Check		OK	06:15	

Data Download Date/Time: 04/26/2018 13:40

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURDOCK,TROY,J
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: MURDOCK,TROY
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 03/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: PATULA

DATE/TIME OF TEST: 04/20/2018 03:28:10

Diagnostics Check		OK	03:21
Air Blank		0.000	03:22
Control Test		0.079	03:22
Air Blank		0.000	03:23
Subject Sample	# 1	0.113	03:23
Breath Volume		2.953	03:23
Air Blank		0.000	03:24
Air Blank		0.000	03:26
Subject Sample	# 2	0.115	03:26
Breath Volume		3.289	03:26
Air Blank		0.000	03:27
Control Test		0.079	03:27
Air Blank		0.000	03:28
Diagnostics Check		OK	03:28

Data Download Date/Time: 04/26/2018 13:40

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,TAYLOR,C
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: ANDERSON,TAYLOR
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 03/28/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: TRACY

DATE/TIME OF TEST: 04/20/2018 23:48:14

Diagnostics Check		OK	23:41
Air Blank		0.000	23:42
Control Test		0.079	23:42
Air Blank		0.000	23:43
Subject Sample	# 1	0.216	23:43
Breath Volume		2.496	23:43
Air Blank		0.000	23:44
Air Blank		0.000	23:46
Subject Sample	# 2	0.207	23:46
Breath Volume		2.300	23:46
Air Blank		0.000	23:47
Control Test		0.079	23:47
Air Blank		0.000	23:48
Diagnostics Check		OK	23:48

Data Download Date/Time: 04/26/2018 13:40

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAYE,MAKINA,A
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: JOHNSON,JAMES
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 03/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: FATIGATE

DATE/TIME OF TEST: 04/26/2018 01:42:24

Diagnostics Check		OK	01:25	
Air Blank		0.000	01:25	
Control Test		0.080	01:26	
Air Blank		0.000	01:26	
Subject Sample	# 1	0.081	01:29	Volume Not Met
Breath Volume		0.390	01:29	
Air Blank		0.000	01:30	
Air Blank		0.000	01:32	
Subject Sample	# 2	0.090	01:35	
Breath Volume		1.148	01:35	
Air Blank		0.000	01:36	
Air Blank		0.000	01:37	
Subject Sample	# 3	0.082	01:40	
Breath Volume		1.105	01:40	
Air Blank		0.000	01:41	
Control Test		0.079	01:41	
Air Blank		0.000	01:42	
Diagnostics Check		OK	01:42	

Data Download Date/Time: 04/26/2018 13:40

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURDOCK,TROY,J
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: MURDOCK,TROY
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 04/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:54

SUBJECT'S LAST NAME: LUEBKE

DATE/TIME OF TEST: 04/28/2018 19:29:10

Diagnostics Check		OK	19:22
Air Blank		0.000	19:22
Control Test		0.079	19:22
Air Blank		0.000	19:23
Subject Sample	# 1	0.395	19:24
Breath Volume		2.812	19:24
Air Blank		0.000	19:25
Air Blank		0.000	19:26
Subject Sample	# 2	0.389	19:27
Breath Volume		2.125	19:27
Air Blank		0.000	19:28
Control Test		0.078	19:28
Air Blank		0.000	19:29
Diagnostics Check		OK	19:29

Data Download Date/Time: 05/24/2018 11:23

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIRANDA,RICHARD,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KARAYIANES,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: HIDALGO

DATE/TIME OF TEST: 04/01/2018 05:01:25

Diagnostics Check		OK	04:53
Air Blank		0.000	04:53
Control Test		0.079	04:54
Air Blank		0.000	04:54
Subject Sample	# 1	0.215	04:55
Breath Volume		1.585	04:55
Air Blank		0.000	04:56
Air Blank		0.000	04:58
Subject Sample	# 2	0.207	04:59
Breath Volume		1.925	04:59
Air Blank		0.000	05:00
Control Test		0.077	05:00
Air Blank		0.000	05:01
Diagnostics Check		OK	05:01

Data Download Date/Time: 04/09/2018 12:40

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,BRENT,D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CLARK,BRENT
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: FALLER

DATE/TIME OF TEST: 04/02/2018 22:10:38

Diagnostics Check		OK	22:04
Air Blank		0.000	22:04
Control Test		0.079	22:04
Air Blank		0.000	22:05
Subject Sample	# 1	0.000	22:05
Breath Volume		1.523	22:05
Air Blank		0.000	22:06
Air Blank		0.000	22:08
Subject Sample	# 2	0.000	22:09
Breath Volume		1.226	22:09
Air Blank		0.000	22:09
Control Test		0.078	22:10
Air Blank		0.000	22:10
Diagnostics Check		OK	22:10

Data Download Date/Time: 04/09/2018 12:40

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,EMILY,K
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WALDO,AARON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: MCLAUGHLIN

DATE/TIME OF TEST: 04/07/2018 00:27:45

Diagnostics Check		OK	00:24	
Air Blank		0.000	00:25	
Control Test		0.078	00:25	
Air Blank		0.000	00:25	
Subject Sample	# 1		00:26	Subject Test Refused
Breath Volume		0.000	00:26	
Air Blank		0.000	00:26	
Control Test		0.078	00:27	
Air Blank		0.000	00:27	
Diagnostics Check		OK	00:27	

Data Download Date/Time: 04/09/2018 12:40

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKINNER,DAVID,D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,K
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:08

SUBJECT'S LAST NAME: MCELROY

DATE/TIME OF TEST: 04/08/2018 19:34:56

Diagnostics Check		OK	19:28
Air Blank		0.000	19:29
Control Test		0.078	19:29
Air Blank		0.000	19:30
Subject Sample	# 1	0.108	19:30
Breath Volume		2.429	19:30
Air Blank		0.000	19:31
Air Blank		0.000	19:32
Subject Sample	# 2	0.102	19:33
Breath Volume		2.621	19:33
Air Blank		0.000	19:34
Control Test		0.078	19:34
Air Blank		0.000	19:34
Diagnostics Check		OK	19:34

Data Download Date/Time: 04/09/2018 12:40

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA,VLADIMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: BURTON

DATE/TIME OF TEST: 04/11/2018 22:45:59

Diagnostics Check		OK	22:43	
Air Blank		0.000	22:43	
Control Test		0.078	22:43	
Air Blank		0.000	22:44	
Subject Sample	# 1		22:44	Subject Test Refused
Breath Volume		0.000	22:44	
Air Blank		0.000	22:45	
Control Test		0.079	22:45	
Air Blank		0.000	22:45	
Diagnostics Check		OK	22:45	

Data Download Date/Time: 05/15/2018 08:33

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,EMILY,K
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KARAYIANES,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 04/13/2018 03:57:08

Diagnostics Check		OK	03:44	
Air Blank		0.000	03:44	
Control Test		0.078	03:44	
Air Blank		0.000	03:45	
Subject Sample	# 1	0.160	03:48	Volume Not Met
Breath Volume		0.179	03:48	
Air Blank		0.000	03:49	
Air Blank		0.000	03:51	
Subject Sample	# 2	0.175	03:51	
Breath Volume		2.160	03:51	
Air Blank		0.000	03:52	
Air Blank		0.000	03:54	
Subject Sample	# 3	0.170	03:55	
Breath Volume		2.769	03:55	
Air Blank		0.000	03:56	
Control Test		0.076	03:56	
Air Blank		0.000	03:57	
Diagnostics Check		OK	03:57	

Data Download Date/Time: 05/15/2018 08:33

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: REFUSE,RONY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 04/13/2018 22:06:39

Diagnostics Check		OK	22:00
Air Blank		0.000	22:00
Control Test		0.078	22:01
Air Blank		0.000	22:01
Subject Sample	# 1	0.226	22:02
Breath Volume		2.167	22:02
Air Blank		0.000	22:02
Air Blank		0.000	22:04
Subject Sample	# 2	0.221	22:05
Breath Volume		2.582	22:05
Air Blank		0.000	22:05
Control Test		0.077	22:06
Air Blank		0.000	22:06
Diagnostics Check		OK	22:06

Data Download Date/Time: 05/15/2018 08:33

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDO,AARON,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA,VLADIMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: BICKEL

DATE/TIME OF TEST: 04/14/2018 22:00:50

Diagnostics Check		OK	21:54
Air Blank		0.000	21:55
Control Test		0.078	21:55
Air Blank		0.000	21:56
Subject Sample	# 1	0.000	21:56
Breath Volume		3.652	21:56
Air Blank		0.000	21:57
Air Blank		0.000	21:59
Subject Sample	# 2	0.000	21:59
Breath Volume		2.980	21:59
Air Blank		0.000	21:59
Control Test		0.078	22:00
Air Blank		0.000	22:00
Diagnostics Check		OK	22:00

Data Download Date/Time: 05/15/2018 08:33

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,BRENT,D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CLARK,BRENT
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: HINZ

DATE/TIME OF TEST: 04/15/2018 22:46:46

Diagnostics Check		OK	22:38
Air Blank		0.000	22:39
Control Test		0.078	22:39
Air Blank		0.000	22:40
Subject Sample	# 1	0.211	22:41
Breath Volume		1.347	22:41
Air Blank		0.000	22:42
Air Blank		0.000	22:44
Subject Sample	# 2	0.214	22:45
Breath Volume		1.707	22:45
Air Blank		0.000	22:45
Control Test		0.077	22:46
Air Blank		0.000	22:46
Diagnostics Check		OK	22:46

Data Download Date/Time: 05/15/2018 08:33

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,EMILY,K
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CRUM,ZACKARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 04/19/2018 04:04:32

Diagnostics Check		OK	04:01	
Air Blank		0.000	04:01	
Control Test		0.079	04:02	
Air Blank		0.000	04:02	
Subject Sample	# 1		04:03	Subject Test Refused
Breath Volume		0.000	04:03	
Air Blank		0.000	04:03	
Control Test		0.079	04:03	
Air Blank		0.000	04:04	
Diagnostics Check		OK	04:04	

Data Download Date/Time: 05/15/2018 08:33

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUM,ZACKARY,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KARAYIANES,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 04/20/2018 23:18:17

Diagnostics Check		OK	23:11
Air Blank		0.000	23:12
Control Test		0.079	23:12
Air Blank		0.000	23:13
Subject Sample	# 1	0.000	23:14
Breath Volume		2.074	23:14
Air Blank		0.000	23:14
Air Blank		0.000	23:16
Subject Sample	# 2	0.000	23:16
Breath Volume		1.867	23:16
Air Blank		0.000	23:17
Control Test		0.079	23:17
Air Blank		0.000	23:18
Diagnostics Check		OK	23:18

Data Download Date/Time: 05/15/2018 08:33

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIRANDA,RICHARD,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SAIA,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:45

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 04/21/2018 07:17:22

Diagnostics Check		OK	07:05	
Air Blank		0.000	07:06	
Control Test		0.079	07:06	
Air Blank		0.000	07:07	
Subject Sample	# 1		07:10	No Sample Provided
Breath Volume		0.000	07:10	
Air Blank		0.000	07:10	
Air Blank		0.000	07:12	
Subject Sample	# 2		07:16	No Sample Provided
Breath Volume		0.000	07:16	
Air Blank		0.000	07:16	
Control Test		0.079	07:16	
Air Blank		0.000	07:17	
Diagnostics Check		OK	07:17	

Data Download Date/Time: 05/15/2018 08:33

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACRI,MICHEAL,N
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ACRI,MICHEAL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 04/25/2018 00:30:06

Diagnostics Check		OK	00:23
Air Blank		0.000	00:24
Control Test		0.078	00:24
Air Blank		0.000	00:24
Subject Sample	# 1	0.190	00:25
Breath Volume		1.710	00:25
Air Blank		0.000	00:26
Air Blank		0.000	00:28
Subject Sample	# 2	0.190	00:28
Breath Volume		2.308	00:28
Air Blank		0.000	00:29
Control Test		0.077	00:29
Air Blank		0.000	00:30
Diagnostics Check		OK	00:30

Data Download Date/Time: 05/15/2018 08:33

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDO,AARON,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SMITH,EMILY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: RAEF

DATE/TIME OF TEST: 04/25/2018 22:45:37

Diagnostics Check		OK	22:38
Air Blank		0.000	22:39
Control Test		0.080	22:39
Air Blank		0.000	22:39
Subject Sample	# 1	0.000	22:40
Breath Volume		1.746	22:40
Air Blank		0.000	22:40
Air Blank		0.000	22:42
Subject Sample	# 2	0.000	22:44
Breath Volume		2.242	22:44
Air Blank		0.000	22:44
Control Test		0.079	22:45
Air Blank		0.000	22:45
Diagnostics Check		OK	22:45

Data Download Date/Time: 05/15/2018 08:33

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,EMILY,K
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WALDO,AARON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:04

SUBJECT'S LAST NAME: BOND

DATE/TIME OF TEST: 04/26/2018 19:31:10

Diagnostics Check		OK	19:24
Air Blank		0.000	19:25
Control Test		0.079	19:25
Air Blank		0.000	19:26
Subject Sample	# 1	0.028	19:26
Breath Volume		3.792	19:26
Air Blank		0.000	19:27
Air Blank		0.000	19:29
Subject Sample	# 2	0.028	19:29
Breath Volume		4.167	19:29
Air Blank		0.000	19:30
Control Test		0.079	19:30
Air Blank		0.000	19:31
Diagnostics Check		OK	19:31

Data Download Date/Time: 05/15/2018 08:33

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,BRENT,D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WALDO,AARON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 04/27/2018 23:56:43

Diagnostics Check		OK	23:53	
Air Blank		0.000	23:53	
Control Test		0.079	23:54	
Air Blank		0.000	23:54	
Subject Sample	# 1		23:55	Subject Test Refused
Breath Volume		0.000	23:55	
Air Blank		0.000	23:55	
Control Test		0.079	23:56	
Air Blank		0.000	23:56	
Diagnostics Check		OK	23:56	

Data Download Date/Time: 05/15/2018 08:33

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CRUM,ZACKARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 04/28/2018 01:11:34

Diagnostics Check		OK	01:05
Air Blank		0.000	01:06
Control Test		0.079	01:06
Air Blank		0.000	01:06
Subject Sample	# 1	0.158	01:07
Breath Volume		2.613	01:07
Air Blank		0.000	01:07
Air Blank		0.000	01:09
Subject Sample	# 2	0.154	01:10
Breath Volume		2.980	01:10
Air Blank		0.000	01:10
Control Test		0.078	01:10
Air Blank		0.000	01:11
Diagnostics Check		OK	01:11

Data Download Date/Time: 05/15/2018 08:34

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUM,ZACKARY,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WALDO,AARON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:48

SUBJECT'S LAST NAME: LOTT

DATE/TIME OF TEST: 04/28/2018 21:14:30

Diagnostics Check		OK	21:08
Air Blank		0.000	21:08
Control Test		0.078	21:09
Air Blank		0.000	21:09
Subject Sample	# 1	0.197	21:10
Breath Volume		1.898	21:10
Air Blank		0.000	21:10
Air Blank		0.000	21:12
Subject Sample	# 2	0.199	21:13
Breath Volume		2.351	21:13
Air Blank		0.000	21:13
Control Test		0.076	21:13
Air Blank		0.000	21:14
Diagnostics Check		OK	21:14

Data Download Date/Time: 05/15/2018 08:34

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDO,AARON,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CRUM,ZACKARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: PHAPASERTH

DATE/TIME OF TEST: 04/28/2018 22:35:19

Diagnostics Check		OK	22:29
Air Blank		0.000	22:29
Control Test		0.078	22:29
Air Blank		0.000	22:30
Subject Sample	# 1	0.115	22:30
Breath Volume		2.960	22:30
Air Blank		0.000	22:31
Air Blank		0.000	22:33
Subject Sample	# 2	0.118	22:33
Breath Volume		1.816	22:33
Air Blank		0.000	22:34
Control Test		0.078	22:34
Air Blank		0.000	22:35
Diagnostics Check		OK	22:35

Data Download Date/Time: 05/15/2018 08:34

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WORTH,KEVIN,J
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: BECERRA,J
ARRESTING OFFICER AGENCY: Lake Placid PD
LAST AGENCY INSPECTION: 03/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: GUZMAN

DATE/TIME OF TEST: 04/01/2018 01:37:19

Diagnostics Check		OK	01:30
Air Blank		0.000	01:30
Control Test		0.080	01:31
Air Blank		0.000	01:31
Subject Sample	# 1	0.167	01:32
Breath Volume		1.960	01:32
Air Blank		0.000	01:33
Air Blank		0.000	01:35
Subject Sample	# 2	0.159	01:35
Breath Volume		1.320	01:35
Air Blank		0.000	01:36
Control Test		0.080	01:36
Air Blank		0.000	01:37
Diagnostics Check		OK	01:37

Data Download Date/Time: 04/25/2018 07:39

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRACKETT,MARK,O
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: BRACKETT,MARK
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 03/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: ZERO TOL
ARREST TIME: 16:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:25

SUBJECT'S LAST NAME: GUZMAN

DATE/TIME OF TEST: 04/02/2018 16:55:34

Diagnostics Check		OK	16:49
Air Blank		0.000	16:49
Control Test		0.080	16:50
Air Blank		0.000	16:50
Subject Sample	# 1	0.000	16:51
Breath Volume		1.851	16:51
Air Blank		0.000	16:51
Air Blank		0.000	16:53
Subject Sample	# 2	0.000	16:54
Breath Volume		3.027	16:54
Air Blank		0.000	16:54
Control Test		0.081	16:54
Air Blank		0.000	16:55
Diagnostics Check		OK	16:55

Data Download Date/Time: 04/25/2018 07:39

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINISH,BARRY,A
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SALVADOR,ELDER
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 03/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: GONZALES

DATE/TIME OF TEST: 04/04/2018 03:13:43

Diagnostics Check		OK	03:07
Air Blank		0.000	03:07
Control Test		0.079	03:08
Air Blank		0.000	03:08
Subject Sample	# 1	0.163	03:09
Breath Volume		2.507	03:09
Air Blank		0.000	03:09
Air Blank		0.000	03:11
Subject Sample	# 2	0.167	03:12
Breath Volume		2.394	03:12
Air Blank		0.000	03:12
Control Test		0.082	03:13
Air Blank		0.000	03:13
Diagnostics Check		OK	03:13

Data Download Date/Time: 04/25/2018 07:39

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROE,ROBERT,D
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: ROE,ROBERT
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 03/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: COURT ORD
ARREST TIME: 16:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: FISHER

DATE/TIME OF TEST: 04/05/2018 16:37:10

Diagnostics Check		OK	16:31
Air Blank		0.000	16:31
Control Test		0.081	16:31
Air Blank		0.000	16:32
Subject Sample	# 1	0.000	16:32
Breath Volume		1.976	16:32
Air Blank		0.000	16:33
Air Blank		0.000	16:35
Subject Sample	# 2	0.000	16:35
Breath Volume		2.031	16:35
Air Blank		0.000	16:36
Control Test		0.080	16:36
Air Blank		0.000	16:37
Diagnostics Check		OK	16:37

Data Download Date/Time: 04/25/2018 07:39

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,SUSAN,L
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: O'MALLEY,M
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 03/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 04/07/2018 03:24:45

Diagnostics Check		OK	03:17
Air Blank		0.000	03:18
Control Test		0.080	03:18
Air Blank		0.000	03:19
Subject Sample	# 1	0.140	03:19
Breath Volume		1.410	03:19
Air Blank		0.000	03:20
Air Blank		0.000	03:22
Subject Sample	# 2	0.139	03:23
Breath Volume		1.375	03:23
Air Blank		0.000	03:23
Control Test		0.080	03:24
Air Blank		0.000	03:24
Diagnostics Check		OK	03:24

Data Download Date/Time: 04/25/2018 07:39

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SALVADOR,ELDER
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 03/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: MILLS

DATE/TIME OF TEST: 04/08/2018 00:49:41

Diagnostics Check		OK	00:43
Air Blank		0.000	00:43
Control Test		0.080	00:43
Air Blank		0.000	00:44
Subject Sample	# 1	0.139	00:45
Breath Volume		1.722	00:45
Air Blank		0.000	00:45
Air Blank		0.000	00:47
Subject Sample	# 2	0.138	00:48
Breath Volume		1.433	00:48
Air Blank		0.000	00:48
Control Test		0.081	00:49
Air Blank		0.000	00:49
Diagnostics Check		OK	00:49

Data Download Date/Time: 04/25/2018 07:39

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GLASCO,GREGORY,S
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SALVADOR,ELDER
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 03/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: CARD

DATE/TIME OF TEST: 04/12/2018 03:28:20

Diagnostics Check		OK	03:22	
Air Blank		0.000	03:23	
Control Test		0.082	03:23	
Air Blank		0.000	03:24	
Subject Sample	# 1	0.040	03:27	
Breath Volume		1.144	03:27	
Air Blank			03:27	Purge Fail
Air Blank		0.000	03:28	

Data Download Date/Time: 04/25/2018 07:39

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SVENDSEN,C,E
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SVENDSEN,C
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 03/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: COURT ORD
ARREST TIME: 14:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: FISHER

DATE/TIME OF TEST: 04/13/2018 14:51:45

Diagnostics Check		OK	14:45
Air Blank		0.000	14:45
Control Test		0.080	14:46
Air Blank		0.000	14:46
Subject Sample	# 1	0.000	14:47
Breath Volume		2.621	14:47
Air Blank		0.000	14:47
Air Blank		0.000	14:49
Subject Sample	# 2	0.000	14:50
Breath Volume		1.699	14:50
Air Blank		0.000	14:50
Control Test		0.083	14:51
Air Blank		0.000	14:51
Diagnostics Check		OK	14:51

Data Download Date/Time: 04/25/2018 07:39

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARLOW,ROYCE,B
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SALVADOR,ELDER
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 03/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 04/21/2018 00:01:15

Diagnostics Check		OK	23:54
Air Blank		0.000	23:55
Control Test		0.080	23:55
Air Blank		0.000	23:55
Subject Sample	# 1	0.255	23:56
Breath Volume		1.347	23:56
Air Blank		0.000	23:57
Air Blank		0.000	23:59
Subject Sample	# 2	0.238	23:59
Breath Volume		1.308	23:59
Air Blank		0.000	00:00
Control Test		0.081	00:00
Air Blank		0.000	00:01
Diagnostics Check		OK	00:01

Data Download Date/Time: 04/25/2018 07:39

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNSED,GLENN,M
BREATH TEST OPERATOR AGENCY: Sebring PD
ARRESTING OFFICER: SERRANO,D
ARRESTING OFFICER AGENCY: Sebring PD
LAST AGENCY INSPECTION: 03/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: ADDIS

DATE/TIME OF TEST: 04/21/2018 01:25:48

Diagnostics Check		OK	01:19
Air Blank		0.000	01:19
Control Test		0.080	01:20
Air Blank		0.000	01:20
Subject Sample	# 1	0.189	01:21
Breath Volume		2.878	01:21
Air Blank		0.000	01:21
Air Blank		0.000	01:23
Subject Sample	# 2	0.175	01:24
Breath Volume		2.230	01:24
Air Blank		0.000	01:24
Control Test		0.080	01:25
Air Blank		0.000	01:25
Diagnostics Check		OK	01:25

Data Download Date/Time: 04/25/2018 07:39

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARLOW,ROYCE,B
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SALVADOR,ELDER
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 03/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: AARON

DATE/TIME OF TEST: 04/22/2018 01:12:57

Diagnostics Check		OK	01:06
Air Blank		0.000	01:07
Control Test		0.081	01:07
Air Blank		0.000	01:08
Subject Sample	# 1	0.179	01:08
Breath Volume		2.460	01:08
Air Blank		0.000	01:09
Air Blank		0.000	01:11
Subject Sample	# 2	0.175	01:11
Breath Volume		2.390	01:11
Air Blank		0.000	01:12
Control Test		0.081	01:12
Air Blank		0.000	01:12
Diagnostics Check		OK	01:12

Data Download Date/Time: 04/25/2018 07:39

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARLOW,ROYCE,B
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: GREEN,JAMES
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 03/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: TINKHAM

DATE/TIME OF TEST: 04/22/2018 03:29:13

Diagnostics Check		OK	03:23
Air Blank		0.000	03:23
Control Test		0.081	03:23
Air Blank		0.000	03:24
Subject Sample	# 1	0.223	03:24
Breath Volume		3.664	03:24
Air Blank		0.000	03:25
Air Blank		0.000	03:27
Subject Sample	# 2	0.224	03:27
Breath Volume		3.937	03:27
Air Blank		0.000	03:28
Control Test		0.081	03:28
Air Blank		0.000	03:29
Diagnostics Check		OK	03:29

Data Download Date/Time: 04/25/2018 07:39

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

INSTRUMENT REGISTERED TO: DAVIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAINES,DAVID,K
BREATH TEST OPERATOR AGENCY: Davie PD
ARRESTING OFFICER: FERNANDEZ,KENNY
ARRESTING OFFICER AGENCY: Davie PD
LAST AGENCY INSPECTION: 03/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:06

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 04/01/2018 22:32:52

Diagnostics Check		OK	22:26
Air Blank		0.000	22:26
Control Test		0.082	22:27
Air Blank		0.000	22:27
Subject Sample	# 1	0.190	22:28
Breath Volume		1.898	22:28
Air Blank		0.000	22:28
Air Blank		0.000	22:30
Subject Sample	# 2	0.192	22:31
Breath Volume		2.273	22:31
Air Blank		0.000	22:31
Control Test		0.081	22:32
Air Blank		0.000	22:32
Diagnostics Check		OK	22:32

Data Download Date/Time: 04/27/2018 22:59

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

INSTRUMENT REGISTERED TO: DAVIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STUEHRENBERG,ERIK,C
BREATH TEST OPERATOR AGENCY: Davie PD
ARRESTING OFFICER: BURNS,CLINT
ARRESTING OFFICER AGENCY: Davie PD
LAST AGENCY INSPECTION: 03/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 04:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:58

SUBJECT'S LAST NAME: BOCANEGRA

DATE/TIME OF TEST: 04/07/2018 05:40:07

Diagnostics Check		OK	05:34
Air Blank		0.000	05:34
Control Test		0.081	05:34
Air Blank		0.000	05:35
Subject Sample	# 1	0.099	05:35
Breath Volume		1.546	05:35
Air Blank		0.000	05:36
Air Blank		0.000	05:38
Subject Sample	# 2	0.099	05:38
Breath Volume		2.039	05:38
Air Blank		0.000	05:39
Control Test		0.079	05:39
Air Blank		0.000	05:40
Diagnostics Check		OK	05:40

Data Download Date/Time: 04/27/2018 22:59

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

INSTRUMENT REGISTERED TO: DAVIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JESSUP,BRIAN,C
BREATH TEST OPERATOR AGENCY: Davie PD
ARRESTING OFFICER: PUENTES,YENDI
ARRESTING OFFICER AGENCY: Davie PD
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: ELVIR AGUILAR

DATE/TIME OF TEST: 04/28/2018 00:27:02

Diagnostics Check		OK	00:20
Air Blank		0.000	00:21
Control Test		0.080	00:21
Air Blank		0.000	00:21
Subject Sample	# 1	0.177	00:22
Breath Volume		2.617	00:22
Air Blank		0.000	00:23
Air Blank		0.000	00:24
Subject Sample	# 2	0.178	00:25
Breath Volume		1.195	00:25
Air Blank		0.000	00:26
Control Test		0.080	00:26
Air Blank		0.000	00:26
Diagnostics Check		OK	00:27

Data Download Date/Time: 05/30/2018 00:17

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULDERRIG,SEAN,F
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: MULDERRIG,SEAN
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 03/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 04/11/2018 00:50:05

Diagnostics Check		OK	00:43
Air Blank		0.000	00:43
Control Test		0.079	00:44
Air Blank		0.000	00:44
Subject Sample	# 1	0.164	00:45
Breath Volume		2.628	00:45
Air Blank		0.000	00:45
Air Blank		0.000	00:47
Subject Sample	# 2	0.162	00:48
Breath Volume		2.207	00:48
Air Blank		0.000	00:49
Control Test		0.079	00:49
Air Blank		0.000	00:50
Diagnostics Check		OK	00:50

Data Download Date/Time: 04/30/2018 10:38

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULDERRIG,SEAN,F
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: MULDERRIG,SEAN
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 03/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: VARGAS

DATE/TIME OF TEST: 04/11/2018 23:12:21

Diagnostics Check		OK	23:09	
Air Blank		0.000	23:09	
Control Test		0.080	23:09	
Air Blank		0.000	23:10	
Subject Sample	# 1		23:10	Subject Test Refused
Breath Volume		0.000	23:10	
Air Blank		0.000	23:11	
Control Test		0.080	23:11	
Air Blank		0.000	23:12	
Diagnostics Check		OK	23:12	

Data Download Date/Time: 04/30/2018 10:38

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUACKENBUSH,BRET,C
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: CARVER,S
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 03/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: HUEWITT

DATE/TIME OF TEST: 04/13/2018 01:50:18

Diagnostics Check		OK	01:42
Air Blank		0.000	01:42
Control Test		0.079	01:43
Air Blank		0.000	01:43
Subject Sample	# 1	0.211	01:44
Breath Volume		1.335	01:44
Air Blank		0.000	01:44
Air Blank		0.000	01:46
Subject Sample	# 2	0.195	01:48
Breath Volume		1.277	01:48
Air Blank		0.000	01:49
Control Test		0.080	01:49
Air Blank		0.000	01:50
Diagnostics Check		OK	01:50

Data Download Date/Time: 04/30/2018 10:38

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULDERRIG,SEAN,F
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: MULDERRIG,SEAN
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 03/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: TAPIA-FELICIANO

DATE/TIME OF TEST: 04/13/2018 05:04:19

Diagnostics Check		OK	04:58
Air Blank		0.000	04:58
Control Test		0.080	04:58
Air Blank		0.000	04:59
Subject Sample	# 1	0.175	04:59
Breath Volume		3.203	04:59
Air Blank		0.000	05:00
Air Blank		0.000	05:02
Subject Sample	# 2	0.172	05:02
Breath Volume		2.808	05:02
Air Blank		0.000	05:03
Control Test		0.079	05:03
Air Blank		0.000	05:04
Diagnostics Check		OK	05:04

Data Download Date/Time: 04/30/2018 10:38

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULDERRIG,SEAN,F
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: MULDERRIG,SEAN
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 03/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: GEISLER

DATE/TIME OF TEST: 04/14/2018 21:41:02

Diagnostics Check		OK	21:38	
Air Blank		0.000	21:38	
Control Test		0.080	21:38	
Air Blank		0.000	21:39	
Subject Sample	# 1		21:39	Subject Test Refused
Breath Volume		0.000	21:39	
Air Blank		0.000	21:40	
Control Test		0.079	21:40	
Air Blank		0.000	21:40	
Diagnostics Check		OK	21:41	

Data Download Date/Time: 04/30/2018 10:38

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULDERRIG,SEAN,F
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: MULDERRIG,SEAN
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 03/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: LEMAY

DATE/TIME OF TEST: 04/19/2018 01:30:03

Diagnostics Check		OK	01:24	
Air Blank		0.000	01:24	
Control Test		0.081	01:25	
Air Blank		0.000	01:25	
Subject Sample	# 1		01:28	Subject Test Refused
Breath Volume		0.000	01:28	
Air Blank		0.000	01:29	
Control Test		0.081	01:29	
Air Blank		0.000	01:29	
Diagnostics Check		OK	01:30	

Data Download Date/Time: 04/30/2018 10:38

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULDERRIG,SEAN,F
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: CUEVAS,HIRAM
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 03/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 04/19/2018 04:09:28

Diagnostics Check		OK	04:02
Air Blank		0.000	04:03
Control Test		0.081	04:03
Air Blank		0.000	04:04
Subject Sample	# 1	0.229	04:04
Breath Volume		2.945	04:04
Air Blank		0.000	04:05
Air Blank		0.000	04:07
Subject Sample	# 2	0.245	04:07
Breath Volume		2.785	04:07
Air Blank		0.000	04:08
Control Test		0.079	04:08
Air Blank		0.000	04:09
Diagnostics Check		OK	04:09

Data Download Date/Time: 04/30/2018 10:38

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULDERRIG,SEAN,F
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: MULDERRIG,SEAN
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 03/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: ALFONSO

DATE/TIME OF TEST: 04/19/2018 22:24:23

Diagnostics Check		OK	22:21	
Air Blank		0.000	22:22	
Control Test		0.080	22:22	
Air Blank		0.000	22:22	
Subject Sample	# 1		22:23	Subject Test Refused
Breath Volume		0.000	22:23	
Air Blank		0.000	22:23	
Control Test		0.080	22:23	
Air Blank		0.000	22:24	
Diagnostics Check		OK	22:24	

Data Download Date/Time: 04/30/2018 10:38

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTLE,CORY,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SUTTLE,CORY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 03/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: SEAY

DATE/TIME OF TEST: 04/28/2018 02:09:31

Diagnostics Check		OK	01:53	
Air Blank		0.000	01:53	
Control Test		0.079	01:54	
Air Blank		0.000	01:54	
Subject Sample	# 1	0.173	01:57	Volume Not Met
Breath Volume		0.050	01:57	
Air Blank		0.000	01:58	
Air Blank		0.000	02:00	
Subject Sample	# 2	0.187	02:02	No .020 agreement
Breath Volume		1.531	02:02	
Air Blank		0.000	02:02	
Air Blank		0.000	02:04	
Subject Sample	# 3	0.163	02:07	No .020 agreement
Breath Volume		1.292	02:07	
Air Blank		0.000	02:08	
Control Test		0.078	02:08	
Air Blank		0.000	02:09	
Diagnostics Check		OK	02:09	

Data Download Date/Time: 04/30/2018 10:38

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTLE,CORY,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SUTTLE,CORY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 03/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: SEAY

DATE/TIME OF TEST: 04/28/2018 02:26:04

Diagnostics Check		OK	02:14	
Air Blank		0.000	02:14	
Control Test		0.078	02:15	
Air Blank		0.000	02:15	
Subject Sample	# 1	0.150	02:18	Volume Not Met
Breath Volume		0.144	02:18	
Air Blank		0.000	02:19	
Air Blank		0.000	02:21	
Subject Sample	# 2	0.160	02:24	Volume Not Met
Breath Volume		0.082	02:24	
Air Blank		0.000	02:25	
Control Test		0.078	02:25	
Air Blank		0.000	02:25	
Diagnostics Check		OK	02:26	

Data Download Date/Time: 04/30/2018 10:38

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULDERRIG,SEAN,F
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: MULDERRIG,SEAN
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 03/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 04/28/2018 23:05:13

Diagnostics Check		OK	23:02	
Air Blank		0.000	23:02	
Control Test		0.080	23:03	
Air Blank		0.000	23:03	
Subject Sample	# 1		23:03	Subject Test Refused
Breath Volume		0.000	23:03	
Air Blank		0.000	23:04	
Control Test		0.079	23:04	
Air Blank		0.000	23:05	
Diagnostics Check		OK	23:05	

Data Download Date/Time: 04/30/2018 10:38

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTLE,CORY,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SUTTLE,CORY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 03/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: ENRIQUEZ

DATE/TIME OF TEST: 04/30/2018 04:06:19

Diagnostics Check		OK	04:00
Air Blank		0.000	04:00
Control Test		0.080	04:01
Air Blank		0.000	04:01
Subject Sample	# 1	0.000	04:02
Breath Volume		2.664	04:02
Air Blank		0.000	04:02
Air Blank		0.000	04:04
Subject Sample	# 2	0.000	04:04
Breath Volume		2.140	04:04
Air Blank		0.000	04:05
Control Test		0.080	04:05
Air Blank		0.000	04:06
Diagnostics Check		OK	04:06

Data Download Date/Time: 04/30/2018 10:38

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JASON,E
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: MORENO,STEPHEN
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 03/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: AUCK

DATE/TIME OF TEST: 04/02/2018 21:32:01

Diagnostics Check		OK	21:29	
Air Blank		0.000	21:29	
Control Test		0.078	21:29	
Air Blank		0.000	21:30	
Subject Sample	# 1		21:30	Subject Test Refused
Breath Volume		0.000	21:30	
Air Blank		0.000	21:31	
Control Test		0.079	21:31	
Air Blank		0.000	21:31	
Diagnostics Check		OK	21:32	

Data Download Date/Time: 04/05/2018 16:05

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,JORGE,L
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: HERNANDEZ,JORGE
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 04/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:11

SUBJECT'S LAST NAME: ORIE

DATE/TIME OF TEST: 04/15/2018 17:40:24

Diagnostics Check		OK	17:34
Air Blank		0.000	17:34
Control Test		0.079	17:34
Air Blank		0.000	17:35
Subject Sample	# 1	0.307	17:35
Breath Volume		1.785	17:35
Air Blank		0.000	17:36
Air Blank		0.000	17:38
Subject Sample	# 2	0.308	17:38
Breath Volume		2.089	17:38
Air Blank		0.000	17:39
Control Test		0.079	17:39
Air Blank		0.000	17:40
Diagnostics Check		OK	17:40

Data Download Date/Time: 05/15/2018 15:55

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

INSTRUMENT REGISTERED TO: WINTER HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOONE,BRANDON,S
BREATH TEST OPERATOR AGENCY: Winter Haven PD
ARRESTING OFFICER: BOONE,BRANDON
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 03/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 04/01/2018 04:49:11

Diagnostics Check		OK	04:46	
Air Blank		0.000	04:46	
Control Test		0.079	04:46	
Air Blank		0.000	04:47	
Subject Sample	# 1		04:47	Subject Test Refused
Breath Volume		0.000	04:47	
Air Blank		0.000	04:48	
Control Test		0.079	04:48	
Air Blank		0.000	04:49	
Diagnostics Check		OK	04:49	

Data Download Date/Time: 04/18/2018 11:41

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

INSTRUMENT REGISTERED TO: WINTER HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOONE,BRANDON,S
BREATH TEST OPERATOR AGENCY: Winter Haven PD
ARRESTING OFFICER: BOONE,BRANDON
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 03/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: AGUILAR

DATE/TIME OF TEST: 04/05/2018 02:45:37

Diagnostics Check		OK	02:38
Air Blank		0.000	02:39
Control Test		0.080	02:39
Air Blank		0.000	02:40
Subject Sample	# 1	0.148	02:40
Breath Volume		2.406	02:40
Air Blank		0.000	02:41
Air Blank		0.000	02:43
Subject Sample	# 2	0.149	02:43
Breath Volume		2.835	02:43
Air Blank		0.000	02:44
Control Test		0.080	02:44
Air Blank		0.000	02:45
Diagnostics Check		OK	02:45

Data Download Date/Time: 04/18/2018 11:41

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

INSTRUMENT REGISTERED TO: WINTER HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALLAWAY,BRADLEY,L
BREATH TEST OPERATOR AGENCY: Winter Haven PD
ARRESTING OFFICER: CALLAWAY,BRADLEY
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 03/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: COURT ORD
ARREST TIME: 08:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:38

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 04/11/2018 09:11:19

Diagnostics Check		OK	09:04
Air Blank		0.000	09:05
Control Test		0.079	09:05
Air Blank		0.000	09:06
Subject Sample	# 1	0.000	09:06
Breath Volume		3.277	09:06
Air Blank		0.000	09:07
Air Blank		0.000	09:09
Subject Sample	# 2	0.000	09:09
Breath Volume		2.769	09:09
Air Blank		0.000	09:10
Control Test		0.078	09:10
Air Blank		0.000	09:11
Diagnostics Check		OK	09:11

Data Download Date/Time: 04/18/2018 11:41

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

INSTRUMENT REGISTERED TO: WINTER HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NICHOLS,KENNETH,D
BREATH TEST OPERATOR AGENCY: Winter Haven PD
ARRESTING OFFICER: NICHOLS,KENNETH
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 03/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: PROB/PAROL
ARREST TIME: 08:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:35

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 04/12/2018 09:05:43

Diagnostics Check		OK	08:59
Air Blank		0.000	08:59
Control Test		0.079	09:00
Air Blank		0.000	09:00
Subject Sample	# 1	0.000	09:01
Breath Volume		2.574	09:01
Air Blank		0.000	09:01
Air Blank		0.000	09:03
Subject Sample	# 2	0.000	09:04
Breath Volume		2.468	09:04
Air Blank		0.000	09:04
Control Test		0.080	09:05
Air Blank		0.000	09:05
Diagnostics Check		OK	09:05

Data Download Date/Time: 04/18/2018 11:41

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

INSTRUMENT REGISTERED TO: WINTER HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANTHAM,CODY,W
BREATH TEST OPERATOR AGENCY: Winter Haven PD
ARRESTING OFFICER: GRANTHAM,CODY
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 03/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: CUBBAGE

DATE/TIME OF TEST: 04/13/2018 03:14:58

Diagnostics Check		OK	03:11	
Air Blank		0.000	03:11	
Control Test		0.078	03:11	
Air Blank		0.000	03:12	
Subject Sample	# 1	0.145	03:13	
Breath Volume		2.882	03:13	
Air Blank			03:14	Purge Fail
Air Blank		0.000	03:14	

Data Download Date/Time: 04/18/2018 11:41

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

INSTRUMENT REGISTERED TO: WINTER HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANTHAM,CODY,W
BREATH TEST OPERATOR AGENCY: Winter Haven PD
ARRESTING OFFICER: GRANTHAM,CODY
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 03/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: CUBBAGE

DATE/TIME OF TEST: 04/13/2018 03:24:46

Diagnostics Check		OK	03:16
Air Blank		0.000	03:17
Control Test		0.078	03:17
Air Blank		0.000	03:18
Subject Sample	# 1	0.133	03:19
Breath Volume		1.742	03:19
Air Blank		0.000	03:20
Air Blank		0.000	03:22
Subject Sample	# 2	0.127	03:23
Breath Volume		2.082	03:23
Air Blank		0.000	03:23
Control Test		0.079	03:24
Air Blank		0.000	03:24
Diagnostics Check		OK	03:24

Data Download Date/Time: 04/18/2018 11:41

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

INSTRUMENT REGISTERED TO: WINTER HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOONE,BRANDON,S
BREATH TEST OPERATOR AGENCY: Winter Haven PD
ARRESTING OFFICER: BOONE,BRANDON
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 03/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:47

SUBJECT'S LAST NAME: RIDER

DATE/TIME OF TEST: 04/13/2018 21:14:49

Diagnostics Check		OK	21:08
Air Blank		0.000	21:08
Control Test		0.079	21:08
Air Blank		0.000	21:09
Subject Sample	# 1	0.190	21:10
Breath Volume		2.769	21:10
Air Blank		0.000	21:10
Air Blank		0.000	21:12
Subject Sample	# 2	0.183	21:13
Breath Volume		2.042	21:13
Air Blank		0.000	21:13
Control Test		0.079	21:14
Air Blank		0.000	21:14
Diagnostics Check		OK	21:14

Data Download Date/Time: 04/18/2018 11:41

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

INSTRUMENT REGISTERED TO: WINTER HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALLAWAY,BRADLEY,L
BREATH TEST OPERATOR AGENCY: Winter Haven PD
ARRESTING OFFICER: CALLAWAY,BRADLEY
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 03/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: COURT ORD
ARREST TIME: 08:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:18

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 04/17/2018 08:49:19

Diagnostics Check		OK	08:42
Air Blank		0.000	08:43
Control Test		0.079	08:43
Air Blank		0.000	08:44
Subject Sample	# 1	0.000	08:44
Breath Volume		2.875	08:44
Air Blank		0.000	08:45
Air Blank		0.000	08:47
Subject Sample	# 2	0.000	08:47
Breath Volume		2.750	08:47
Air Blank		0.000	08:48
Control Test		0.080	08:48
Air Blank		0.000	08:49
Diagnostics Check		OK	08:49

Data Download Date/Time: 04/18/2018 11:41

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

INSTRUMENT REGISTERED TO: WINTER HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEBSON,BRET,E
BREATH TEST OPERATOR AGENCY: Winter Haven PD
ARRESTING OFFICER: HEBSON,BRET
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 03/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: COURT ORD
ARREST TIME: 08:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:44

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 04/18/2018 09:21:01

Diagnostics Check		OK	09:14
Air Blank		0.000	09:15
Control Test		0.079	09:15
Air Blank		0.000	09:16
Subject Sample	# 1	0.000	09:16
Breath Volume		2.937	09:16
Air Blank		0.000	09:17
Air Blank		0.000	09:19
Subject Sample	# 2	0.000	09:19
Breath Volume		2.683	09:19
Air Blank		0.000	09:20
Control Test		0.080	09:20
Air Blank		0.000	09:20
Diagnostics Check		OK	09:21

Data Download Date/Time: 04/18/2018 11:41

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

INSTRUMENT REGISTERED TO: WINTER HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOONE,BRANDON,S
BREATH TEST OPERATOR AGENCY: Winter Haven PD
ARRESTING OFFICER: BOONE,BRANDON
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 04/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:38

SUBJECT'S LAST NAME: FELKER

DATE/TIME OF TEST: 04/28/2018 07:08:17

Diagnostics Check		OK	07:01
Air Blank		0.000	07:02
Control Test		0.079	07:02
Air Blank		0.000	07:03
Subject Sample	# 1	0.053	07:03
Breath Volume		2.367	07:03
Air Blank		0.000	07:04
Air Blank		0.000	07:06
Subject Sample	# 2	0.052	07:06
Breath Volume		2.421	07:06
Air Blank		0.000	07:07
Control Test		0.081	07:07
Air Blank		0.000	07:08
Diagnostics Check		OK	07:08

Data Download Date/Time: 05/29/2018 10:58

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

INSTRUMENT REGISTERED TO: GULFPORT PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,TIMOTHY,R
BREATH TEST OPERATOR AGENCY: Gulfport PD
ARRESTING OFFICER: CARTER,TIMOTHY
ARRESTING OFFICER AGENCY: Gulfport PD
LAST AGENCY INSPECTION: 03/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: BEATTY

DATE/TIME OF TEST: 04/12/2018 04:00:04

Diagnostics Check		OK	03:47	
Air Blank		0.000	03:47	
Control Test		0.081	03:47	
Air Blank		0.000	03:48	
Subject Sample	# 1		03:51	No Sample Provided
Breath Volume		0.000	03:51	
Air Blank		0.000	03:52	
Air Blank		0.000	03:54	
Subject Sample	# 2	0.183	03:55	
Breath Volume		1.964	03:55	
Air Blank		0.000	03:55	
Air Blank		0.000	03:57	
Subject Sample	# 3	0.180	03:58	
Breath Volume		2.820	03:58	
Air Blank		0.000	03:59	
Control Test		0.080	03:59	
Air Blank		0.000	03:59	
Diagnostics Check		OK	04:00	

Data Download Date/Time: 04/17/2018 05:11

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

INSTRUMENT REGISTERED TO: GULFPORT PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,TIMOTHY,R
BREATH TEST OPERATOR AGENCY: Gulfport PD
ARRESTING OFFICER: CARTER,TIMOTHY
ARRESTING OFFICER AGENCY: Gulfport PD
LAST AGENCY INSPECTION: 03/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: JANUARY

DATE/TIME OF TEST: 04/14/2018 23:37:32

Diagnostics Check		OK	23:30
Air Blank		0.000	23:31
Control Test		0.079	23:31
Air Blank		0.000	23:32
Subject Sample	# 1	0.216	23:32
Breath Volume		4.289	23:32
Air Blank		0.000	23:33
Air Blank		0.000	23:35
Subject Sample	# 2	0.222	23:36
Breath Volume		2.964	23:36
Air Blank		0.000	23:36
Control Test		0.081	23:36
Air Blank		0.000	23:37
Diagnostics Check		OK	23:37

Data Download Date/Time: 04/17/2018 05:11

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

INSTRUMENT REGISTERED TO: GULFPORT PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,TIMOTHY,R
BREATH TEST OPERATOR AGENCY: Gulfport PD
ARRESTING OFFICER: CARTER,TIMOTHY
ARRESTING OFFICER AGENCY: Gulfport PD
LAST AGENCY INSPECTION: 03/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:34

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 04/15/2018 22:02:09

Diagnostics Check		OK	21:55
Air Blank		0.000	21:55
Control Test		0.080	21:55
Air Blank		0.000	21:56
Subject Sample	# 1	0.263	21:57
Breath Volume		4.136	21:57
Air Blank		0.000	21:57
Air Blank		0.000	21:59
Subject Sample	# 2	0.255	22:00
Breath Volume		2.253	22:00
Air Blank		0.000	22:01
Control Test		0.080	22:01
Air Blank		0.000	22:02
Diagnostics Check		OK	22:02

Data Download Date/Time: 04/17/2018 05:11

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001062

INSTRUMENT REGISTERED TO: TREASURE ISLAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PREMRU,KIP,E
BREATH TEST OPERATOR AGENCY: Treasure Island PD
ARRESTING OFFICER: JANIK,STEPHEN
ARRESTING OFFICER AGENCY: Indian Shores PD
LAST AGENCY INSPECTION: 03/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 04/02/2018 22:06:56

Diagnostics Check		OK	22:00
Air Blank		0.000	22:00
Control Test		0.079	22:01
Air Blank		0.000	22:01
Subject Sample	# 1	0.185	22:02
Breath Volume		1.796	22:02
Air Blank		0.000	22:03
Air Blank		0.000	22:04
Subject Sample	# 2	0.186	22:05
Breath Volume		1.691	22:05
Air Blank		0.000	22:06
Control Test		0.078	22:06
Air Blank		0.000	22:06
Diagnostics Check		OK	22:06

Data Download Date/Time: 04/27/2018 11:50

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001062

INSTRUMENT REGISTERED TO: TREASURE ISLAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PREMRU,KIP,E
BREATH TEST OPERATOR AGENCY: Treasure Island PD
ARRESTING OFFICER: PREMRU,KIP
ARRESTING OFFICER AGENCY: Treasure Island PD
LAST AGENCY INSPECTION: 03/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:17

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 04/23/2018 18:58:49

Diagnostics Check		OK	18:47	
Air Blank		0.000	18:47	
Control Test		0.079	18:47	
Air Blank		0.000	18:48	
Subject Sample	# 1	0.131	18:51	Volume Not Met
Breath Volume		0.207	18:51	
Air Blank		0.000	18:52	
Air Blank		0.000	18:53	
Subject Sample	# 2	0.106	18:57	Volume Not Met
Breath Volume		0.292	18:57	
Air Blank		0.000	18:57	
Control Test		0.078	18:58	
Air Blank		0.000	18:58	
Diagnostics Check		OK	18:58	

Data Download Date/Time: 04/27/2018 11:50

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

INSTRUMENT REGISTERED TO: HOLLYWOOD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BALUJA,JONATHAN,
BREATH TEST OPERATOR AGENCY: Hollywood PD
ARRESTING OFFICER: BALUJA,JONATHAN
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 03/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 06:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:12

SUBJECT'S LAST NAME: GUADALUPE

DATE/TIME OF TEST: 04/01/2018 07:39:35

Diagnostics Check		OK	07:32
Air Blank		0.000	07:33
Control Test		0.076	07:33
Air Blank		0.000	07:33
Subject Sample	# 1	0.169	07:34
Breath Volume		1.167	07:34
Air Blank		0.000	07:35
Air Blank		0.000	07:37
Subject Sample	# 2	0.172	07:38
Breath Volume		2.511	07:38
Air Blank		0.000	07:38
Control Test		0.075	07:38
Air Blank		0.000	07:39
Diagnostics Check		OK	07:39

Data Download Date/Time: 04/28/2018 11:21

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAUGHERTY,DANIEL,M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: DAUGHERTY,DANIEL
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 03/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: BUI
ARREST TIME: 16:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:10

SUBJECT'S LAST NAME: BARNARD

DATE/TIME OF TEST: 04/06/2018 17:38:06

Diagnostics Check		OK	17:30
Air Blank		0.000	17:31
Control Test		0.079	17:31
Air Blank		0.000	17:32
Subject Sample	# 1	0.190	17:32
Breath Volume		1.367	17:32
Air Blank		0.000	17:33
Air Blank		0.000	17:35
Subject Sample	# 2	0.171	17:36
Breath Volume		1.363	17:36
Air Blank		0.000	17:37
Control Test		0.078	17:37
Air Blank		0.000	17:38
Diagnostics Check		OK	17:38

Data Download Date/Time: 04/18/2018 01:19

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ENGEL,JASON,S
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: HAZIME,AL-HUSSEIN
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 03/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: BOBB

DATE/TIME OF TEST: 04/09/2018 04:15:40

Diagnostics Check		OK	04:08
Air Blank		0.000	04:09
Control Test		0.079	04:09
Air Blank		0.000	04:10
Subject Sample	# 1	0.123	04:11
Breath Volume		2.285	04:11
Air Blank		0.000	04:11
Air Blank		0.000	04:13
Subject Sample	# 2	0.127	04:14
Breath Volume		2.250	04:14
Air Blank		0.000	04:14
Control Test		0.078	04:15
Air Blank		0.000	04:15
Diagnostics Check		OK	04:15

Data Download Date/Time: 04/18/2018 01:19

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIBSON,JOHN,A
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: GIBSON,JOHN
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A3
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:25

SUBJECT'S LAST NAME: BROCK

DATE/TIME OF TEST: 04/04/2018 17:54:35

Diagnostics Check		OK	17:48
Air Blank		0.000	17:48
Control Test		0.079	17:48
Air Blank		0.000	17:49
Subject Sample	# 1	0.214	17:50
Breath Volume		2.832	17:50
Air Blank		0.000	17:50
Air Blank		0.000	17:52
Subject Sample	# 2	0.222	17:53
Breath Volume		2.367	17:53
Air Blank		0.000	17:53
Control Test		0.079	17:53
Air Blank		0.000	17:54
Diagnostics Check		OK	17:54

Data Download Date/Time: 04/28/2018 10:56

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRINIDAD,MICHAEL,E
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: WARWICK,ANDY
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A3
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:42

SUBJECT'S LAST NAME: MIRANDA

DATE/TIME OF TEST: 04/11/2018 19:14:33

Diagnostics Check		OK	19:06	
Air Blank		0.000	19:06	
Control Test		0.079	19:06	
Air Blank		0.000	19:07	
Subject Sample	# 1		19:10	No Sample Provided
Breath Volume		0.000	19:10	
Air Blank		0.000	19:10	
Air Blank		0.000	19:12	
Subject Sample	# 2		19:13	Subject Test Refused
Breath Volume		0.000	19:13	
Air Blank		0.000	19:13	
Control Test		0.079	19:13	
Air Blank		0.000	19:14	
Diagnostics Check		OK	19:14	

Data Download Date/Time: 04/28/2018 10:56

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DEERMAN,M
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: BETTIG

DATE/TIME OF TEST: 04/01/2018 00:40:22

Diagnostics Check		OK	00:26	
Air Blank		0.000	00:27	
Control Test		0.079	00:27	
Air Blank		0.000	00:28	
Subject Sample	# 1		00:31	No Sample Provided
Breath Volume		0.000	00:31	
Air Blank		0.000	00:31	
Air Blank		0.000	00:33	
Subject Sample	# 2	0.166	00:34	
Breath Volume		1.179	00:34	
Air Blank		0.000	00:35	
Air Blank		0.000	00:37	
Subject Sample	# 3	0.161	00:38	
Breath Volume		1.363	00:38	
Air Blank		0.000	00:39	
Control Test		0.084	00:39	
Air Blank		0.000	00:40	
Diagnostics Check		OK	00:40	

Data Download Date/Time: 04/27/2018 23:35

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DONOHUE,E
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: RIVERA MAYAS

DATE/TIME OF TEST: 04/01/2018 02:56:57

Diagnostics Check		OK	02:53	
Air Blank		0.000	02:54	
Control Test		0.079	02:54	
Air Blank		0.000	02:54	
Subject Sample	# 1		02:55	Subject Test Refused
Breath Volume		0.000	02:55	
Air Blank		0.000	02:56	
Control Test		0.079	02:56	
Air Blank		0.000	02:56	
Diagnostics Check		OK	02:56	

Data Download Date/Time: 04/27/2018 23:35

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: JOHNSON,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 04/07/2018 05:24:25

Diagnostics Check		OK	05:15
Air Blank		0.000	05:16
Control Test		0.080	05:16
Air Blank		0.000	05:17
Subject Sample	# 1	0.140	05:19
Breath Volume		1.445	05:19
Air Blank		0.000	05:19
Air Blank		0.000	05:21
Subject Sample	# 2	0.143	05:22
Breath Volume		2.394	05:22
Air Blank		0.000	05:23
Control Test		0.081	05:23
Air Blank		0.000	05:24
Diagnostics Check		OK	05:24

Data Download Date/Time: 04/27/2018 23:35

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MASON,G
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: HALEY

DATE/TIME OF TEST: 04/14/2018 04:28:17

Diagnostics Check		OK	04:21
Air Blank		0.000	04:21
Control Test		0.079	04:22
Air Blank		0.000	04:22
Subject Sample	# 1	0.170	04:23
Breath Volume		1.476	04:23
Air Blank		0.000	04:24
Air Blank		0.000	04:26
Subject Sample	# 2	0.169	04:26
Breath Volume		1.500	04:26
Air Blank		0.000	04:27
Control Test		0.079	04:27
Air Blank		0.000	04:28
Diagnostics Check		OK	04:28

Data Download Date/Time: 04/27/2018 23:35

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MARTIN,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: MENES

DATE/TIME OF TEST: 04/19/2018 04:43:16

Diagnostics Check		OK	04:36
Air Blank		0.000	04:37
Control Test		0.079	04:37
Air Blank		0.000	04:38
Subject Sample	# 1	0.167	04:38
Breath Volume		3.339	04:38
Air Blank		0.000	04:39
Air Blank		0.000	04:41
Subject Sample	# 2	0.158	04:41
Breath Volume		2.839	04:41
Air Blank		0.000	04:42
Control Test		0.081	04:42
Air Blank		0.000	04:43
Diagnostics Check		OK	04:43

Data Download Date/Time: 04/27/2018 23:35

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTHONY,MICHAEL,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI,JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: GULA

DATE/TIME OF TEST: 04/22/2018 02:32:21

Diagnostics Check		OK	02:29	
Air Blank		0.000	02:29	
Control Test		0.078	02:30	
Air Blank		0.000	02:30	
Subject Sample	# 1		02:30	Subject Test Refused
Breath Volume		0.000	02:30	
Air Blank		0.000	02:31	
Control Test		0.078	02:31	
Air Blank		0.000	02:32	
Diagnostics Check		OK	02:32	

Data Download Date/Time: 04/27/2018 23:35

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTHONY,MICHAEL,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: WEIGNER

DATE/TIME OF TEST: 04/22/2018 02:43:01

Diagnostics Check		OK	02:36
Air Blank		0.000	02:37
Control Test		0.078	02:37
Air Blank		0.000	02:38
Subject Sample	# 1	0.046	02:38
Breath Volume		3.351	02:38
Air Blank		0.000	02:39
Air Blank		0.000	02:41
Subject Sample	# 2	0.046	02:41
Breath Volume		3.339	02:41
Air Blank		0.000	02:42
Control Test		0.083	02:42
Air Blank		0.000	02:42
Diagnostics Check		OK	02:43

Data Download Date/Time: 04/27/2018 23:35

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 04/28/2018 02:53:56

Diagnostics Check		OK	02:51	
Air Blank		0.000	02:51	
Control Test		0.080	02:51	
Air Blank		0.000	02:52	
Subject Sample	# 1		02:52	Subject Test Refused
Breath Volume		0.000	02:52	
Air Blank		0.000	02:53	
Control Test		0.080	02:53	
Air Blank		0.000	02:53	
Diagnostics Check		OK	02:53	

Data Download Date/Time: 05/31/2018 00:07

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AKINS,G
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: COLEMAN

DATE/TIME OF TEST: 04/29/2018 04:46:46

Diagnostics Check		OK	04:38	
Air Blank		0.000	04:38	
Control Test		0.079	04:39	
Air Blank		0.000	04:39	
Subject Sample	# 1		04:42	No Sample Provided
Breath Volume		0.000	04:42	
Air Blank		0.000	04:43	
Air Blank		0.000	04:45	
Subject Sample	# 2		04:45	Subject Test Refused
Breath Volume		0.000	04:45	
Air Blank		0.000	04:45	
Control Test		0.079	04:46	
Air Blank		0.000	04:46	
Diagnostics Check		OK	04:46	

Data Download Date/Time: 05/31/2018 00:07

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FITHIAN,JASON
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 04/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 05:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:33

SUBJECT'S LAST NAME: MUNOZ RIVERA

DATE/TIME OF TEST: 04/27/2018 06:04:31

Diagnostics Check		OK	05:54
Air Blank		0.000	05:54
Control Test		0.078	05:55
Air Blank		0.000	05:55
Subject Sample	# 1	0.264	05:56
Breath Volume		1.636	05:56
Air Blank		0.000	05:57
Air Blank		0.000	05:59
Subject Sample	# 2	0.285	05:59
Breath Volume		1.558	05:59
Air Blank		0.000	06:00
Air Blank		0.000	06:02
Subject Sample	# 3	0.265	06:03
Breath Volume		1.355	06:03
Air Blank		0.000	06:03
Control Test		0.077	06:03
Air Blank		0.000	06:04
Diagnostics Check		OK	06:04

Data Download Date/Time: 05/10/2018 12:04

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 04/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 04/28/2018 04:10:29

Diagnostics Check		OK	04:04
Air Blank		0.000	04:04
Control Test		0.079	04:04
Air Blank		0.000	04:05
Subject Sample	# 1	0.145	04:05
Breath Volume		2.968	04:05
Air Blank		0.000	04:06
Air Blank		0.000	04:08
Subject Sample	# 2	0.149	04:09
Breath Volume		3.656	04:09
Air Blank		0.000	04:09
Control Test		0.078	04:09
Air Blank		0.000	04:10
Diagnostics Check		OK	04:10

Data Download Date/Time: 05/10/2018 12:04

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: JOHNSON,EDWIN
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 04/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 21:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: SINGH

DATE/TIME OF TEST: 04/28/2018 22:20:38

Diagnostics Check		OK	22:14
Air Blank		0.000	22:15
Control Test		0.078	22:15
Air Blank		0.000	22:15
Subject Sample	# 1	0.161	22:16
Breath Volume		1.253	22:16
Air Blank		0.000	22:16
Air Blank		0.000	22:18
Subject Sample	# 2	0.163	22:19
Breath Volume		1.675	22:19
Air Blank		0.000	22:19
Control Test		0.078	22:20
Air Blank		0.000	22:20
Diagnostics Check		OK	22:20

Data Download Date/Time: 05/10/2018 12:04

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LOUGHRAN,KEVIN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 04/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 04/29/2018 04:10:44

Diagnostics Check		OK	04:04
Air Blank		0.000	04:04
Control Test		0.078	04:05
Air Blank		0.000	04:05
Subject Sample	# 1	0.329	04:06
Breath Volume		1.953	04:06
Air Blank		0.000	04:06
Air Blank		0.000	04:08
Subject Sample	# 2	0.329	04:09
Breath Volume		1.355	04:09
Air Blank		0.000	04:09
Control Test		0.078	04:10
Air Blank		0.000	04:10
Diagnostics Check		OK	04:10

Data Download Date/Time: 05/10/2018 12:04

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ROJAS,RANDY
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 04/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 04/29/2018 04:56:46

Diagnostics Check		OK	04:47
Air Blank		0.000	04:47
Control Test		0.078	04:48
Air Blank		0.000	04:48
Subject Sample	# 1	0.177	04:51
Breath Volume		1.312	04:51
Air Blank		0.000	04:52
Air Blank		0.000	04:53
Subject Sample	# 2	0.185	04:55
Breath Volume		1.886	04:55
Air Blank		0.000	04:55
Control Test		0.078	04:56
Air Blank		0.000	04:56
Diagnostics Check		OK	04:56

Data Download Date/Time: 05/10/2018 12:04

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BOHM,JONATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 04/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 04:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 04/29/2018 05:34:26

Diagnostics Check		OK	05:26
Air Blank		0.000	05:27
Control Test		0.078	05:27
Air Blank		0.000	05:28
Subject Sample	# 1	0.196	05:29
Breath Volume		2.093	05:29
Air Blank		0.000	05:30
Air Blank		0.000	05:31
Subject Sample	# 2	0.191	05:32
Breath Volume		1.777	05:32
Air Blank		0.000	05:33
Control Test		0.077	05:33
Air Blank		0.000	05:34
Diagnostics Check		OK	05:34

Data Download Date/Time: 05/10/2018 12:04

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: STRZALKOWSKI,SEAN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 04/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 04:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:04

SUBJECT'S LAST NAME: SEJAS

DATE/TIME OF TEST: 04/30/2018 05:35:25

Diagnostics Check		OK	05:25
Air Blank		0.000	05:26
Control Test		0.079	05:26
Air Blank		0.000	05:27
Subject Sample	# 1	0.146	05:28
Breath Volume		1.785	05:28
Air Blank		0.000	05:29
Air Blank		0.000	05:30
Subject Sample	# 2	0.146	05:33
Breath Volume		2.683	05:33
Air Blank		0.000	05:34
Control Test		0.079	05:34
Air Blank		0.000	05:35
Diagnostics Check		OK	05:35

Data Download Date/Time: 05/10/2018 12:04

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,CHAD,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,DANIEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 03/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: RAYMOND

DATE/TIME OF TEST: 04/06/2018 03:42:15

Diagnostics Check		OK	03:30	
Air Blank		0.000	03:31	
Control Test		0.080	03:31	
Air Blank		0.000	03:32	
Subject Sample	# 1		03:35	No Sample Provided
Breath Volume		0.000	03:35	
Air Blank		0.000	03:35	
Air Blank		0.000	03:37	
Subject Sample	# 2		03:40	No Sample Provided
Breath Volume		0.000	03:40	
Air Blank		0.000	03:41	
Control Test		0.079	03:41	
Air Blank		0.000	03:42	
Diagnostics Check		OK	03:42	

Data Download Date/Time: 04/24/2018 01:58

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,CHAD,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,DANIEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 03/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 12:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:01

SUBJECT'S LAST NAME: EPPS

DATE/TIME OF TEST: 04/13/2018 12:29:35

Diagnostics Check		OK	12:22
Air Blank		0.000	12:23
Control Test		0.079	12:23
Air Blank		0.000	12:23
Subject Sample	# 1	0.000	12:24
Breath Volume		2.070	12:24
Air Blank		0.000	12:25
Air Blank		0.000	12:27
Subject Sample	# 2	0.000	12:28
Breath Volume		1.871	12:28
Air Blank		0.000	12:28
Control Test		0.079	12:28
Air Blank		0.000	12:29
Diagnostics Check		OK	12:29

Data Download Date/Time: 04/24/2018 01:58

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,CHAD,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: HURT,JONATHAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 03/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: MARTINEZ CALVA

DATE/TIME OF TEST: 04/13/2018 21:59:01

Diagnostics Check		OK	21:53	
Air Blank		0.000	21:53	
Control Test		0.079	21:54	
Air Blank		0.000	21:54	
Subject Sample	# 1	0.158	21:56	
Breath Volume		2.074	21:56	
Air Blank		0.000	21:56	
Air Blank			21:58	Ambient Fail
Air Blank		0.000	21:59	

Data Download Date/Time: 04/24/2018 01:58

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,CHAD,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: HURT,JONATHAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 03/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: MARTINEZ CALVA

DATE/TIME OF TEST: 04/13/2018 22:07:40

Diagnostics Check		OK	22:00
Air Blank		0.000	22:00
Control Test		0.079	22:01
Air Blank		0.000	22:01
Subject Sample	# 1	0.158	22:02
Breath Volume		1.621	22:02
Air Blank		0.000	22:03
Air Blank		0.000	22:04
Subject Sample	# 2	0.163	22:05
Breath Volume		1.687	22:05
Air Blank		0.000	22:06
Control Test		0.079	22:07
Air Blank		0.000	22:07
Diagnostics Check		OK	22:07

Data Download Date/Time: 04/24/2018 01:58

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,CHAD,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,DANIEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 03/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: STRABLE

DATE/TIME OF TEST: 04/19/2018 19:53:01

Diagnostics Check		OK	19:46
Air Blank		0.000	19:46
Control Test		0.080	19:46
Air Blank		0.000	19:47
Subject Sample	# 1	0.000	19:48
Breath Volume		3.894	19:48
Air Blank		0.000	19:49
Air Blank		0.000	19:51
Subject Sample	# 2	0.000	19:51
Breath Volume		3.828	19:51
Air Blank		0.000	19:52
Control Test		0.077	19:52
Air Blank		0.000	19:52
Diagnostics Check		OK	19:53

Data Download Date/Time: 04/24/2018 01:58

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,CHAD,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: HURT,JONATHAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 03/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: MORDEN

DATE/TIME OF TEST: 04/22/2018 01:14:18

Diagnostics Check		OK	01:07
Air Blank		0.000	01:07
Control Test		0.079	01:07
Air Blank		0.000	01:08
Subject Sample	# 1	0.162	01:09
Breath Volume		2.328	01:09
Air Blank		0.000	01:10
Air Blank		0.000	01:11
Subject Sample	# 2	0.153	01:12
Breath Volume		3.062	01:12
Air Blank		0.000	01:13
Control Test		0.078	01:13
Air Blank		0.000	01:14
Diagnostics Check		OK	01:14

Data Download Date/Time: 04/24/2018 01:58

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUGENT,DAVID,A
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NUGENT,DAVID
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: RAMIREZ OLIVARES

DATE/TIME OF TEST: 04/28/2018 02:07:32

Diagnostics Check		OK	02:00
Air Blank		0.000	02:01
Control Test		0.079	02:01
Air Blank		0.000	02:01
Subject Sample	# 1	0.097	02:02
Breath Volume		1.808	02:02
Air Blank		0.000	02:03
Air Blank		0.000	02:05
Subject Sample	# 2	0.095	02:05
Breath Volume		1.832	02:05
Air Blank		0.000	02:06
Control Test		0.079	02:06
Air Blank		0.000	02:07
Diagnostics Check		OK	02:07

Data Download Date/Time: 05/03/2018 18:47

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VENEZIA,STEVEN,L
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: VENEZIA,STEVEN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: ACUNA-HERNANDEZ

DATE/TIME OF TEST: 04/28/2018 03:06:51

Diagnostics Check		OK	03:03	
Air Blank		0.000	03:03	
Control Test		0.079	03:04	
Air Blank		0.000	03:04	
Subject Sample	# 1	0.169	03:05	
Breath Volume		2.199	03:05	
Air Blank			03:06	Purge Fail
Air Blank		0.000	03:06	

Data Download Date/Time: 05/03/2018 18:47

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VENEZIA,STEVEN,L
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: VENEZIA,STEVEN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: ACUNA-HERNANDEZ

DATE/TIME OF TEST: 04/28/2018 03:11:37

Diagnostics Check	OK	03:10	
Air Blank	0.000	03:10	
Control Test	0.070	03:10	Control Outside Tolerance
Air Blank	0.000	03:11	

Data Download Date/Time: 05/03/2018 18:47

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VENEZIA,STEVEN,L
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: VENEZIA,STEVEN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: ACUNA-HERNANDEZ

DATE/TIME OF TEST: 04/28/2018 03:20:55

Diagnostics Check		OK	03:13
Air Blank		0.000	03:14
Control Test		0.078	03:14
Air Blank		0.000	03:15
Subject Sample	# 1	0.173	03:15
Breath Volume		1.437	03:15
Air Blank		0.000	03:16
Air Blank		0.000	03:18
Subject Sample	# 2	0.181	03:19
Breath Volume		1.179	03:19
Air Blank		0.000	03:19
Control Test		0.077	03:20
Air Blank		0.000	03:20
Diagnostics Check		OK	03:20

Data Download Date/Time: 05/03/2018 18:47

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEGERSMITH,DANIEL,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,DANIEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: OLAZARAN-VILLAFUERTE

DATE/TIME OF TEST: 04/28/2018 20:28:57

Diagnostics Check		OK	20:24	
Air Blank		0.000	20:24	
Control Test		0.080	20:25	
Air Blank		0.000	20:25	
Subject Sample	# 1	0.190	20:27	
Breath Volume		1.679	20:27	
Air Blank			20:28	Purge Fail
Air Blank		0.000	20:28	

Data Download Date/Time: 05/03/2018 18:47

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEGERSMITH,DANIEL,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,DANIEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: OLAZARAN-VILLAFUERTE

DATE/TIME OF TEST: 04/28/2018 20:40:22

Diagnostics Check		OK	20:33
Air Blank		0.000	20:33
Control Test		0.078	20:33
Air Blank		0.000	20:34
Subject Sample	# 1	0.188	20:35
Breath Volume		2.097	20:35
Air Blank		0.000	20:36
Air Blank		0.000	20:37
Subject Sample	# 2	0.186	20:38
Breath Volume		1.867	20:38
Air Blank		0.000	20:39
Control Test		0.077	20:39
Air Blank		0.000	20:40
Diagnostics Check		OK	20:40

Data Download Date/Time: 05/03/2018 18:47

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICHMOND,MATTHEW,D
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: MOCCIA,JAY
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 08:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:20

SUBJECT'S LAST NAME: NOPAL CRUZ

DATE/TIME OF TEST: 04/29/2018 08:45:14

Diagnostics Check		OK	08:41	
Air Blank		0.000	08:41	
Control Test		0.080	08:41	
Air Blank		0.000	08:42	
Subject Sample	# 1		08:43	Subject Test Refused
Breath Volume		0.000	08:43	
Air Blank		0.000	08:44	
Control Test		0.081	08:44	
Air Blank		0.000	08:45	
Diagnostics Check		OK	08:45	

Data Download Date/Time: 05/03/2018 18:47

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OBRIEN,DANIEL,P
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: GIORDANO,NIC
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: KHAYKHAM

DATE/TIME OF TEST: 04/30/2018 01:39:22

Diagnostics Check		OK	01:31
Air Blank		0.000	01:31
Control Test		0.080	01:32
Air Blank		0.000	01:32
Subject Sample	# 1	0.203	01:34
Breath Volume		1.132	01:34
Air Blank		0.000	01:35
Air Blank		0.000	01:36
Subject Sample	# 2	0.191	01:37
Breath Volume		1.417	01:37
Air Blank		0.000	01:38
Control Test		0.079	01:38
Air Blank		0.000	01:39
Diagnostics Check		OK	01:39

Data Download Date/Time: 05/03/2018 18:47

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARNES,BRYA,J
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: KARNES,B
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 10:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:12

SUBJECT'S LAST NAME: KOSCIOLEK

DATE/TIME OF TEST: 04/30/2018 11:40:25

Diagnostics Check		OK	11:33
Air Blank		0.000	11:34
Control Test		0.079	11:34
Air Blank		0.000	11:35
Subject Sample	# 1	0.226	11:35
Breath Volume		2.371	11:35
Air Blank		0.000	11:36
Air Blank		0.000	11:38
Subject Sample	# 2	0.220	11:38
Breath Volume		2.730	11:38
Air Blank		0.000	11:39
Control Test		0.079	11:39
Air Blank		0.000	11:40
Diagnostics Check		OK	11:40

Data Download Date/Time: 05/03/2018 18:47

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,DAKOTA,E
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 04/22/2018 03:37:58

Diagnostics Check		OK	03:31
Air Blank		0.000	03:31
Control Test		0.079	03:32
Air Blank		0.000	03:32
Subject Sample	# 1	0.155	03:33
Breath Volume		1.750	03:33
Air Blank		0.000	03:33
Air Blank		0.000	03:35
Subject Sample	# 2	0.150	03:36
Breath Volume		2.054	03:36
Air Blank		0.000	03:37
Control Test		0.078	03:37
Air Blank		0.000	03:37
Diagnostics Check		OK	03:37

Data Download Date/Time: 05/19/2018 14:36

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNOW,CHRIS,D
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: CORDERIO,E
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: VAUGHNER

DATE/TIME OF TEST: 04/29/2018 00:21:52

Diagnostics Check		OK	00:10	
Air Blank		0.000	00:10	
Control Test		0.080	00:11	
Air Blank		0.000	00:11	
Subject Sample	# 1	0.187	00:14	Volume Not Met
Breath Volume		0.050	00:14	
Air Blank		0.000	00:15	
Air Blank		0.000	00:17	
Subject Sample	# 2		00:20	No Sample Provided
Breath Volume		0.000	00:20	
Air Blank		0.000	00:20	
Control Test		0.078	00:21	
Air Blank		0.000	00:21	
Diagnostics Check		OK	00:21	

Data Download Date/Time: 05/19/2018 14:36

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNOW,CHRIS,D
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JOHNSON,G
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: VETTRAINO

DATE/TIME OF TEST: 04/29/2018 01:56:26

Diagnostics Check		OK	01:50
Air Blank		0.000	01:50
Control Test		0.079	01:51
Air Blank		0.000	01:51
Subject Sample	# 1	0.147	01:52
Breath Volume		1.726	01:52
Air Blank		0.000	01:52
Air Blank		0.000	01:54
Subject Sample	# 2	0.141	01:54
Breath Volume		1.433	01:54
Air Blank		0.000	01:55
Control Test		0.079	01:55
Air Blank		0.000	01:56
Diagnostics Check		OK	01:56

Data Download Date/Time: 05/19/2018 14:36

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAEZ HERNANDEZ,DENYS,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: PAEZ HERNANDEZ,DENYS
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: CHRISTIAN

DATE/TIME OF TEST: 04/01/2018 04:57:26

Diagnostics Check		OK	04:50
Air Blank		0.000	04:51
Control Test		0.078	04:51
Air Blank		0.000	04:52
Subject Sample	# 1	0.187	04:52
Breath Volume		2.789	04:52
Air Blank		0.000	04:53
Air Blank		0.000	04:55
Subject Sample	# 2	0.187	04:55
Breath Volume		2.417	04:55
Air Blank		0.000	04:56
Control Test		0.078	04:56
Air Blank		0.000	04:57
Diagnostics Check		OK	04:57

Data Download Date/Time: 04/25/2018 14:19

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,RAEFORD,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GRIFFIN,RAEFORD
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: D\ALESSANDRO

DATE/TIME OF TEST: 04/03/2018 00:36:16

Diagnostics Check		OK	00:30
Air Blank		0.000	00:30
Control Test		0.080	00:31
Air Blank		0.000	00:31
Subject Sample	# 1	0.000	00:32
Breath Volume		3.851	00:32
Air Blank		0.000	00:32
Air Blank		0.000	00:34
Subject Sample	# 2	0.000	00:34
Breath Volume		2.921	00:34
Air Blank		0.000	00:35
Control Test		0.080	00:35
Air Blank		0.000	00:36
Diagnostics Check		OK	00:36

Data Download Date/Time: 04/25/2018 14:19

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRIOLAT,CASEY,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BRIOLAT,CASEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: GRUERIO

DATE/TIME OF TEST: 04/03/2018 23:55:04

Diagnostics Check		OK	23:46	
Air Blank		0.000	23:46	
Control Test		0.078	23:47	
Air Blank		0.000	23:47	
Subject Sample	# 1		23:50	No Sample Provided
Breath Volume		0.000	23:50	
Air Blank		0.000	23:51	
Air Blank		0.000	23:53	
Subject Sample	# 2		23:53	Subject Test Refused
Breath Volume		0.000	23:53	
Air Blank		0.000	23:54	
Control Test		0.079	23:54	
Air Blank		0.000	23:54	
Diagnostics Check		OK	23:55	

Data Download Date/Time: 04/25/2018 14:19

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,BARRY,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: NIXON,BARRY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 03/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:34

SUBJECT'S LAST NAME: ARMBRISTER

DATE/TIME OF TEST: 04/07/2018 20:01:06

Diagnostics Check		OK	19:54
Air Blank		0.000	19:55
Control Test		0.079	19:55
Air Blank		0.000	19:56
Subject Sample	# 1	0.192	19:56
Breath Volume		2.894	19:56
Air Blank		0.000	19:57
Air Blank		0.000	19:59
Subject Sample	# 2	0.183	19:59
Breath Volume		2.320	19:59
Air Blank		0.000	20:00
Control Test		0.078	20:00
Air Blank		0.000	20:01
Diagnostics Check		OK	20:01

Data Download Date/Time: 04/25/2018 14:19

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,BARRY,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: NIXON,BARRY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 03/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: JAMIESON

DATE/TIME OF TEST: 04/08/2018 03:22:47

Diagnostics Check		OK	03:16
Air Blank		0.000	03:16
Control Test		0.079	03:17
Air Blank		0.000	03:17
Subject Sample	# 1	0.116	03:18
Breath Volume		1.738	03:18
Air Blank		0.000	03:18
Air Blank		0.000	03:20
Subject Sample	# 2	0.100	03:21
Breath Volume		2.144	03:21
Air Blank		0.000	03:21
Control Test		0.078	03:22
Air Blank		0.000	03:22
Diagnostics Check		OK	03:22

Data Download Date/Time: 04/25/2018 14:19

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVARADO,ANGEL,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ALVARADO,ANGEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:56

SUBJECT'S LAST NAME: PHELPS

DATE/TIME OF TEST: 04/08/2018 05:22:56

Diagnostics Check		OK	05:16
Air Blank		0.000	05:16
Control Test		0.078	05:17
Air Blank		0.000	05:17
Subject Sample	# 1	0.189	05:18
Breath Volume		1.390	05:18
Air Blank		0.000	05:18
Air Blank		0.000	05:20
Subject Sample	# 2	0.176	05:21
Breath Volume		1.171	05:21
Air Blank		0.000	05:22
Control Test		0.078	05:22
Air Blank		0.000	05:22
Diagnostics Check		OK	05:22

Data Download Date/Time: 04/25/2018 14:19

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,RAEFORD,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GRIFFIN,RAEFORD
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: LUCKEY

DATE/TIME OF TEST: 04/08/2018 20:56:45

Diagnostics Check		OK	20:45	
Air Blank		0.000	20:45	
Control Test		0.079	20:46	
Air Blank		0.000	20:46	
Subject Sample	# 1		20:49	No Sample Provided
Breath Volume		0.000	20:49	
Air Blank		0.000	20:50	
Air Blank		0.000	20:52	
Subject Sample	# 2		20:55	No Sample Provided
Breath Volume		0.000	20:55	
Air Blank		0.000	20:55	
Control Test		0.078	20:56	
Air Blank		0.000	20:56	
Diagnostics Check		OK	20:56	

Data Download Date/Time: 04/25/2018 14:19

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,ADAM,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: MORRIS,ADAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: ALGOR

DATE/TIME OF TEST: 04/13/2018 05:30:21

Diagnostics Check		OK	05:26	
Air Blank		0.000	05:26	
Control Test		0.079	05:26	
Air Blank		0.000	05:27	
Subject Sample	# 1		05:28	Subject Test Refused
Breath Volume		0.000	05:28	
Air Blank		0.000	05:29	
Control Test		0.079	05:29	
Air Blank		0.000	05:30	
Diagnostics Check		OK	05:30	

Data Download Date/Time: 04/25/2018 14:19

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,BARRY,M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: NIXON,BARRY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 03/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: STEINBRENNER

DATE/TIME OF TEST: 04/13/2018 22:25:43

Diagnostics Check		OK	22:19
Air Blank		0.000	22:19
Control Test		0.079	22:20
Air Blank		0.000	22:20
Subject Sample	# 1	0.254	22:21
Breath Volume		2.234	22:21
Air Blank		0.000	22:21
Air Blank		0.000	22:23
Subject Sample	# 2	0.262	22:24
Breath Volume		2.656	22:24
Air Blank		0.000	22:24
Control Test		0.079	22:25
Air Blank		0.000	22:25
Diagnostics Check		OK	22:25

Data Download Date/Time: 04/25/2018 14:19

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINTERS,LANE,A
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: HAMMELMAN,T
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 03/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: LAMB

DATE/TIME OF TEST: 04/17/2018 03:01:39

Diagnostics Check		OK	02:55
Air Blank		0.000	02:55
Control Test		0.078	02:55
Air Blank		0.000	02:56
Subject Sample	# 1	0.181	02:57
Breath Volume		3.625	02:57
Air Blank		0.000	02:57
Air Blank		0.000	02:59
Subject Sample	# 2	0.182	03:00
Breath Volume		3.281	03:00
Air Blank		0.000	03:00
Control Test		0.078	03:01
Air Blank		0.000	03:01
Diagnostics Check		OK	03:01

Data Download Date/Time: 04/25/2018 14:19

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,BARRY,M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: NIXON,BARRY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 03/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: AYRES

DATE/TIME OF TEST: 04/21/2018 02:15:14

Diagnostics Check		OK	02:09
Air Blank		0.000	02:09
Control Test		0.078	02:09
Air Blank		0.000	02:10
Subject Sample	# 1	0.184	02:10
Breath Volume		2.488	02:10
Air Blank		0.000	02:11
Air Blank		0.000	02:13
Subject Sample	# 2	0.172	02:13
Breath Volume		1.410	02:13
Air Blank		0.000	02:14
Control Test		0.079	02:14
Air Blank		0.000	02:15
Diagnostics Check		OK	02:15

Data Download Date/Time: 04/25/2018 14:19

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUSE,RONNIE,L
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: JOHNSON,CIERRA
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 03/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 04/01/2018 00:14:27

Diagnostics Check		OK	00:06
Air Blank		0.000	00:06
Control Test		0.077	00:07
Air Blank		0.000	00:07
Subject Sample	# 1	0.191	00:09
Breath Volume		2.234	00:09
Air Blank		0.000	00:10
Air Blank		0.000	00:12
Subject Sample	# 2	0.195	00:12
Breath Volume		2.054	00:12
Air Blank		0.000	00:13
Control Test		0.076	00:13
Air Blank		0.000	00:14
Diagnostics Check		OK	00:14

Data Download Date/Time: 04/27/2018 16:26

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASSICK,PATRICIA,A
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: AONAYA,PETER
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 11:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 04/09/2018 02:38:18

Diagnostics Check		OK	02:30
Air Blank		0.000	02:31
Control Test		0.077	02:31
Air Blank		0.000	02:32
Subject Sample	# 1	0.108	02:33
Breath Volume		1.164	02:33
Air Blank		0.000	02:34
Air Blank		0.000	02:36
Subject Sample	# 2	0.102	02:36
Breath Volume		1.359	02:36
Air Blank		0.000	02:37
Control Test		0.076	02:37
Air Blank		0.000	02:38
Diagnostics Check		OK	02:38

Data Download Date/Time: 04/27/2018 16:26

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANDA,ROBERT,W
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: ABBOTT,RYAN
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 03/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 14:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:08

SUBJECT'S LAST NAME: MOLINA DE LA CRUZ

DATE/TIME OF TEST: 04/09/2018 15:31:41

Diagnostics Check		OK	15:28	
Air Blank		0.000	15:29	
Control Test		0.076	15:29	
Air Blank		0.000	15:30	
Subject Sample	# 1		15:30	Subject Test Refused
Breath Volume		0.000	15:30	
Air Blank		0.000	15:30	
Control Test		0.077	15:31	
Air Blank		0.000	15:31	
Diagnostics Check		OK	15:31	

Data Download Date/Time: 04/27/2018 16:26

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASSICK,PATRICIA,A
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: WARD,SHANE
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 03/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:23

SUBJECT'S LAST NAME: MORALES-GUTIERREZ

DATE/TIME OF TEST: 04/22/2018 20:52:28

Diagnostics Check		OK	20:44
Air Blank		0.000	20:44
Control Test		0.077	20:44
Air Blank		0.000	20:45
Subject Sample	# 1	0.251	20:46
Breath Volume		1.835	20:46
Air Blank		0.000	20:47
Air Blank		0.000	20:49
Subject Sample	# 2	0.237	20:50
Breath Volume		1.222	20:50
Air Blank		0.000	20:51
Control Test		0.075	20:51
Air Blank		0.000	20:52
Diagnostics Check		OK	20:52

Data Download Date/Time: 04/27/2018 16:26

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUSE,RONNIE,L
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: GICKER,CHRISTOPHER
ARRESTING OFFICER AGENCY: Wauchula PD
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: LOPEZ SANTIAGO

DATE/TIME OF TEST: 04/29/2018 00:37:54

Diagnostics Check		OK	00:31
Air Blank		0.000	00:31
Control Test		0.077	00:32
Air Blank		0.000	00:32
Subject Sample	# 1	0.188	00:33
Breath Volume		1.843	00:33
Air Blank		0.000	00:33
Air Blank		0.000	00:35
Subject Sample	# 2	0.190	00:36
Breath Volume		1.386	00:36
Air Blank		0.000	00:36
Control Test		0.077	00:37
Air Blank		0.000	00:37
Diagnostics Check		OK	00:37

Data Download Date/Time: 05/29/2018 11:44

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WALDO,AARON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: DANLEY

DATE/TIME OF TEST: 04/06/2018 03:53:33

Diagnostics Check		OK	03:47
Air Blank		0.000	03:47
Control Test		0.080	03:47
Air Blank		0.000	03:48
Subject Sample	# 1	0.089	03:49
Breath Volume		1.687	03:49
Air Blank		0.000	03:49
Air Blank		0.000	03:51
Subject Sample	# 2	0.080	03:51
Breath Volume		2.070	03:51
Air Blank		0.000	03:52
Control Test		0.080	03:52
Air Blank		0.000	03:53
Diagnostics Check		OK	03:53

Data Download Date/Time: 04/09/2018 12:46

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CRUM,ZACKARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: SCARBOROUGH

DATE/TIME OF TEST: 04/06/2018 20:06:02

Diagnostics Check		OK	19:59
Air Blank		0.000	19:59
Control Test		0.080	19:59
Air Blank		0.000	20:00
Subject Sample	# 1	0.239	20:00
Breath Volume		2.226	20:00
Air Blank		0.000	20:01
Air Blank		0.000	20:03
Subject Sample	# 2	0.238	20:04
Breath Volume		2.203	20:04
Air Blank		0.000	20:05
Control Test		0.080	20:05
Air Blank		0.000	20:05
Diagnostics Check		OK	20:06

Data Download Date/Time: 04/09/2018 12:46

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SMITH,EMILY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:37

SUBJECT'S LAST NAME: HELTON

DATE/TIME OF TEST: 04/06/2018 22:09:12

Diagnostics Check		OK	22:02
Air Blank		0.000	22:03
Control Test		0.080	22:03
Air Blank		0.000	22:04
Subject Sample	# 1	0.259	22:04
Breath Volume		3.230	22:04
Air Blank		0.000	22:05
Air Blank		0.000	22:07
Subject Sample	# 2	0.252	22:07
Breath Volume		3.042	22:07
Air Blank		0.000	22:08
Control Test		0.079	22:08
Air Blank		0.000	22:09
Diagnostics Check		OK	22:09

Data Download Date/Time: 04/09/2018 12:46

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUM,ZACKARY,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA,VLADAMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: KASPER

DATE/TIME OF TEST: 04/06/2018 23:35:14

Diagnostics Check		OK	23:28
Air Blank		0.000	23:29
Control Test		0.079	23:29
Air Blank		0.000	23:30
Subject Sample	# 1	0.182	23:30
Breath Volume		2.996	23:30
Air Blank		0.000	23:31
Air Blank		0.000	23:33
Subject Sample	# 2	0.180	23:33
Breath Volume		2.925	23:33
Air Blank		0.000	23:34
Control Test		0.079	23:34
Air Blank		0.000	23:35
Diagnostics Check		OK	23:35

Data Download Date/Time: 04/09/2018 12:46

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUM,ZACKARY,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WALDO,AARON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: AMONS

DATE/TIME OF TEST: 04/07/2018 02:31:26

Diagnostics Check		OK	02:28	
Air Blank		0.000	02:28	
Control Test		0.081	02:28	
Air Blank		0.000	02:29	
Subject Sample	# 1		02:30	Subject Test Refused
Breath Volume		0.000	02:30	
Air Blank		0.000	02:30	
Control Test		0.080	02:30	
Air Blank		0.000	02:31	
Diagnostics Check		OK	02:31	

Data Download Date/Time: 04/09/2018 12:46

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUM,ZACKARY,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA,VLADAMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: AUCOIN

DATE/TIME OF TEST: 04/07/2018 03:41:41

Diagnostics Check		OK	03:35
Air Blank		0.000	03:35
Control Test		0.080	03:36
Air Blank		0.000	03:36
Subject Sample	# 1	0.148	03:37
Breath Volume		3.750	03:37
Air Blank		0.000	03:37
Air Blank		0.000	03:39
Subject Sample	# 2	0.136	03:40
Breath Volume		3.812	03:40
Air Blank		0.000	03:40
Control Test		0.080	03:41
Air Blank		0.000	03:41
Diagnostics Check		OK	03:41

Data Download Date/Time: 04/09/2018 12:46

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDO,AARON,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CRUM,ZACKARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:02

SUBJECT'S LAST NAME: SKARR

DATE/TIME OF TEST: 04/07/2018 20:32:03

Diagnostics Check		OK	20:28	
Air Blank		0.000	20:29	
Control Test		0.081	20:29	
Air Blank		0.000	20:30	
Subject Sample	# 1		20:30	Subject Test Refused
Breath Volume		0.000	20:30	
Air Blank		0.000	20:31	
Control Test		0.080	20:31	
Air Blank		0.000	20:31	
Diagnostics Check		OK	20:32	

Data Download Date/Time: 04/09/2018 12:46

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SMITH,EMILY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: MUNVISETH

DATE/TIME OF TEST: 04/14/2018 02:25:00

Diagnostics Check		OK	02:17
Air Blank		0.000	02:18
Control Test		0.080	02:18
Air Blank		0.000	02:18
Subject Sample	# 1	0.125	02:20
Breath Volume		1.207	02:20
Air Blank		0.000	02:20
Air Blank		0.000	02:22
Subject Sample	# 2	0.119	02:23
Breath Volume		1.285	02:23
Air Blank		0.000	02:24
Control Test		0.077	02:24
Air Blank		0.000	02:24
Diagnostics Check		OK	02:25

Data Download Date/Time: 05/15/2018 08:36

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CRUM,ZACHARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: STROUSE

DATE/TIME OF TEST: 04/14/2018 03:59:40

Diagnostics Check		OK	03:56	
Air Blank		0.000	03:57	
Control Test		0.080	03:57	
Air Blank		0.000	03:57	
Subject Sample	# 1		03:58	Subject Test Refused
Breath Volume		0.000	03:58	
Air Blank		0.000	03:58	
Control Test		0.080	03:59	
Air Blank		0.000	03:59	
Diagnostics Check		OK	03:59	

Data Download Date/Time: 05/15/2018 08:36

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUM,ZACKARY,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WALDO,AARON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: LUCIANO VICENTE

DATE/TIME OF TEST: 04/19/2018 01:31:45

Diagnostics Check		OK	01:28	
Air Blank		0.000	01:28	
Control Test		0.080	01:29	
Air Blank		0.000	01:29	
Subject Sample	# 1		01:30	Subject Test Refused
Breath Volume		0.000	01:30	
Air Blank		0.000	01:30	
Control Test		0.081	01:31	
Air Blank		0.000	01:31	
Diagnostics Check		OK	01:31	

Data Download Date/Time: 05/15/2018 08:36

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SMITH,EMILY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:47

SUBJECT'S LAST NAME: OCASIO

DATE/TIME OF TEST: 04/19/2018 22:23:12

Diagnostics Check		OK	22:16
Air Blank		0.000	22:17
Control Test		0.081	22:17
Air Blank		0.000	22:17
Subject Sample	# 1	0.185	22:18
Breath Volume		1.832	22:18
Air Blank		0.000	22:19
Air Blank		0.000	22:20
Subject Sample	# 2	0.195	22:21
Breath Volume		2.726	22:21
Air Blank		0.000	22:22
Control Test		0.079	22:22
Air Blank		0.000	22:23
Diagnostics Check		OK	22:23

Data Download Date/Time: 05/15/2018 08:37

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUM,ZACKARY,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA,VLADAMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 04/20/2018 01:59:25

Diagnostics Check		OK	01:53
Air Blank		0.000	01:53
Control Test		0.080	01:54
Air Blank		0.000	01:54
Subject Sample	# 1	0.174	01:55
Breath Volume		1.171	01:55
Air Blank		0.000	01:55
Air Blank		0.000	01:57
Subject Sample	# 2	0.172	01:57
Breath Volume		1.753	01:57
Air Blank		0.000	01:58
Control Test		0.080	01:58
Air Blank		0.000	01:59
Diagnostics Check		OK	01:59

Data Download Date/Time: 05/15/2018 08:37

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CRUM,ZACKARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: ST LUCE

DATE/TIME OF TEST: 04/21/2018 00:00:37

Diagnostics Check		OK	23:57	
Air Blank		0.000	23:58	
Control Test		0.080	23:58	
Air Blank		0.000	23:59	
Subject Sample	# 1		23:59	Subject Test Refused
Breath Volume		0.000	23:59	
Air Blank		0.000	23:59	
Control Test		0.080	00:00	
Air Blank		0.000	00:00	
Diagnostics Check		OK	00:00	

Data Download Date/Time: 05/15/2018 08:37

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KARAYIANES,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: DENTON

DATE/TIME OF TEST: 04/21/2018 01:44:35

Diagnostics Check		OK	01:41	
Air Blank		0.000	01:42	
Control Test		0.080	01:42	
Air Blank		0.000	01:42	
Subject Sample	# 1		01:43	Subject Test Refused
Breath Volume		0.000	01:43	
Air Blank		0.000	01:43	
Control Test		0.079	01:43	
Air Blank		0.000	01:44	
Diagnostics Check		OK	01:44	

Data Download Date/Time: 05/15/2018 08:37

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUM,ZACKARY,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA,VLADIMER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:16

SUBJECT'S LAST NAME: PACHECO

DATE/TIME OF TEST: 04/25/2018 21:51:52

Diagnostics Check		OK	21:47	
Air Blank		0.000	21:48	
Control Test		0.080	21:48	
Air Blank		0.000	21:49	
Subject Sample	# 1		21:50	Subject Test Refused
Breath Volume		0.000	21:50	
Air Blank		0.000	21:50	
Control Test		0.080	21:51	
Air Blank		0.000	21:51	
Diagnostics Check		OK	21:51	

Data Download Date/Time: 05/15/2018 08:37

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CRUM,ZACKARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: FASON

DATE/TIME OF TEST: 04/27/2018 00:32:03

Diagnostics Check		OK	00:26
Air Blank		0.000	00:26
Control Test		0.081	00:26
Air Blank		0.000	00:27
Subject Sample	# 1	0.000	00:27
Breath Volume		1.933	00:27
Air Blank		0.000	00:28
Air Blank		0.000	00:30
Subject Sample	# 2	0.000	00:30
Breath Volume		1.968	00:30
Air Blank		0.000	00:31
Control Test		0.080	00:31
Air Blank		0.000	00:31
Diagnostics Check		OK	00:32

Data Download Date/Time: 05/15/2018 08:37

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUM,ZACKARY,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA,VLADIMER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 04/27/2018 23:31:46

Diagnostics Check		OK	23:23
Air Blank		0.000	23:24
Control Test		0.080	23:24
Air Blank		0.000	23:24
Subject Sample	# 1	0.219	23:26
Breath Volume		1.410	23:26
Air Blank		0.000	23:26
Air Blank		0.000	23:28
Subject Sample	# 2	0.227	23:30
Breath Volume		1.445	23:30
Air Blank		0.000	23:30
Control Test		0.081	23:31
Air Blank		0.000	23:31
Diagnostics Check		OK	23:31

Data Download Date/Time: 05/15/2018 08:37

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WALDO,AARON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: JALOOT JALOOT

DATE/TIME OF TEST: 04/28/2018 02:50:08

Diagnostics Check		OK	02:41
Air Blank		0.000	02:41
Control Test		0.081	02:41
Air Blank		0.000	02:42
Subject Sample	# 1	0.042	02:45
Breath Volume		2.335	02:45
Air Blank		0.000	02:45
Air Blank		0.000	02:47
Subject Sample	# 2	0.041	02:48
Breath Volume		1.558	02:48
Air Blank		0.000	02:49
Control Test		0.079	02:49
Air Blank		0.000	02:50
Diagnostics Check		OK	02:50

Data Download Date/Time: 05/15/2018 08:37

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA,VLADIMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: SILVESTRI

DATE/TIME OF TEST: 04/28/2018 21:51:36

Diagnostics Check		OK	21:48	
Air Blank		0.000	21:49	
Control Test		0.079	21:49	
Air Blank		0.000	21:50	
Subject Sample	# 1		21:50	Subject Test Refused
Breath Volume		0.000	21:50	
Air Blank		0.000	21:50	
Control Test		0.080	21:51	
Air Blank		0.000	21:51	
Diagnostics Check		OK	21:51	

Data Download Date/Time: 05/15/2018 08:37

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WALDO,AARON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: CRISWELL

DATE/TIME OF TEST: 04/29/2018 02:03:38

Diagnostics Check		OK	01:55
Air Blank		0.000	01:56
Control Test		0.080	01:56
Air Blank		0.000	01:57
Subject Sample	# 1	0.141	01:59
Breath Volume		2.472	01:59
Air Blank		0.000	01:59
Air Blank		0.000	02:01
Subject Sample	# 2	0.129	02:02
Breath Volume		1.242	02:02
Air Blank		0.000	02:02
Control Test		0.078	02:03
Air Blank		0.000	02:03
Diagnostics Check		OK	02:03

Data Download Date/Time: 05/15/2018 08:37

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUM,ZACKARY,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA,VLADIMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 04/29/2018 03:03:05

Diagnostics Check		OK	02:57	
Air Blank		0.000	02:58	
Control Test		0.080	02:58	
Air Blank		0.000	02:59	
Subject Sample	# 1		03:01	Subject Test Refused
Breath Volume		0.000	03:01	
Air Blank		0.000	03:02	
Control Test		0.080	03:02	
Air Blank		0.000	03:03	
Diagnostics Check		OK	03:03	

Data Download Date/Time: 05/15/2018 08:37

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASTINGS,KENNETH,S
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: HASTINGS,KENNETH
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 03/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: IOVANNI

DATE/TIME OF TEST: 04/01/2018 01:46:14

Diagnostics Check		OK	01:43	
Air Blank		0.000	01:43	
Control Test		0.080	01:44	
Air Blank		0.000	01:44	
Subject Sample	# 1		01:44	Subject Test Refused
Breath Volume		0.000	01:44	
Air Blank		0.000	01:45	
Control Test		0.080	01:45	
Air Blank		0.000	01:46	
Diagnostics Check		OK	01:46	

Data Download Date/Time: 04/11/2018 12:38

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 03/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 04/01/2018 03:34:04

Diagnostics Check		OK	03:30	
Air Blank		0.000	03:31	
Control Test		0.080	03:31	
Air Blank		0.000	03:31	
Subject Sample	# 1		03:32	Subject Test Refused
Breath Volume		0.000	03:32	
Air Blank		0.000	03:33	
Control Test		0.080	03:33	
Air Blank		0.000	03:33	
Diagnostics Check		OK	03:34	

Data Download Date/Time: 04/11/2018 12:38

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TINDALL,RICHARD,P
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: TINDALL,RICHARD
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 03/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: WALTERS

DATE/TIME OF TEST: 04/02/2018 06:10:30

Diagnostics Check		OK	06:00
Air Blank		0.000	06:01
Control Test		0.079	06:01
Air Blank		0.000	06:02
Subject Sample	# 1	0.193	06:02
Breath Volume		2.089	06:02
Air Blank		0.000	06:03
Air Blank		0.000	06:05
Subject Sample	# 2	0.216	06:05
Breath Volume		1.972	06:05
Air Blank		0.000	06:06
Air Blank		0.000	06:08
Subject Sample	# 3	0.205	06:08
Breath Volume		1.933	06:08
Air Blank		0.000	06:09
Control Test		0.079	06:09
Air Blank		0.000	06:10
Diagnostics Check		OK	06:10

Data Download Date/Time: 04/11/2018 12:38

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICE,JOHN,B
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: RICE,JOHN
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 03/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 04/02/2018 23:03:11

Diagnostics Check		OK	22:56
Air Blank		0.000	22:56
Control Test		0.080	22:57
Air Blank		0.000	22:57
Subject Sample	# 1	0.130	22:58
Breath Volume		2.796	22:58
Air Blank		0.000	22:59
Air Blank		0.000	23:01
Subject Sample	# 2	0.132	23:01
Breath Volume		2.851	23:01
Air Blank		0.000	23:02
Control Test		0.080	23:02
Air Blank		0.000	23:03
Diagnostics Check		OK	23:03

Data Download Date/Time: 04/11/2018 12:38

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 03/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: WILCOX

DATE/TIME OF TEST: 04/07/2018 00:27:56

Diagnostics Check		OK	00:19
Air Blank		0.000	00:20
Control Test		0.080	00:20
Air Blank		0.000	00:21
Subject Sample	# 1	0.187	00:23
Breath Volume		2.246	00:23
Air Blank		0.000	00:23
Air Blank		0.000	00:25
Subject Sample	# 2	0.189	00:26
Breath Volume		2.320	00:26
Air Blank		0.000	00:27
Control Test		0.080	00:27
Air Blank		0.000	00:27
Diagnostics Check		OK	00:27

Data Download Date/Time: 04/11/2018 12:38

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 03/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: LAMBERT

DATE/TIME OF TEST: 04/07/2018 03:08:26

Diagnostics Check		OK	02:59
Air Blank		0.000	03:00
Control Test		0.081	03:00
Air Blank		0.000	03:01
Subject Sample	# 1	0.217	03:03
Breath Volume		1.750	03:03
Air Blank		0.000	03:03
Air Blank		0.000	03:05
Subject Sample	# 2	0.226	03:06
Breath Volume		1.738	03:06
Air Blank		0.000	03:07
Control Test		0.080	03:07
Air Blank		0.000	03:08
Diagnostics Check		OK	03:08

Data Download Date/Time: 04/11/2018 12:38

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASTINGS,KENNETH,S
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: HASTINGS,KENNETH
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 03/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: PRYOR

DATE/TIME OF TEST: 04/08/2018 01:58:55

Diagnostics Check		OK	01:52
Air Blank		0.000	01:53
Control Test		0.080	01:53
Air Blank		0.000	01:53
Subject Sample	# 1	0.189	01:54
Breath Volume		1.339	01:54
Air Blank		0.000	01:54
Air Blank		0.000	01:56
Subject Sample	# 2	0.183	01:57
Breath Volume		1.453	01:57
Air Blank		0.000	01:58
Control Test		0.079	01:58
Air Blank		0.000	01:58
Diagnostics Check		OK	01:58

Data Download Date/Time: 04/11/2018 12:38

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUNT,JERRY,D
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: HUNT,JERRY
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 04/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 13:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:57

SUBJECT'S LAST NAME: GORE

DATE/TIME OF TEST: 04/12/2018 14:23:37

Diagnostics Check		OK	14:20	
Air Blank		0.000	14:20	
Control Test		0.080	14:21	
Air Blank		0.000	14:21	
Subject Sample	# 1		14:22	Subject Test Refused
Breath Volume		0.000	14:22	
Air Blank		0.000	14:22	
Control Test		0.081	14:23	
Air Blank		0.000	14:23	
Diagnostics Check		OK	14:23	

Data Download Date/Time: 05/29/2018 11:15

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASTINGS,KENNETH,S
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: HASTINGS,KENNETH
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 04/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: FAULK

DATE/TIME OF TEST: 04/13/2018 22:58:02

Diagnostics Check		OK	22:50
Air Blank		0.000	22:50
Control Test		0.079	22:51
Air Blank		0.000	22:51
Subject Sample	# 1	0.088	22:53
Breath Volume		1.355	22:53
Air Blank		0.000	22:54
Air Blank		0.000	22:56
Subject Sample	# 2	0.087	22:56
Breath Volume		1.351	22:56
Air Blank		0.000	22:57
Control Test		0.080	22:57
Air Blank		0.000	22:57
Diagnostics Check		OK	22:58

Data Download Date/Time: 05/29/2018 11:15

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASTINGS,KENNETH,S
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: HASTINGS,KENNETH
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 04/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: JOHNSTON

DATE/TIME OF TEST: 04/19/2018 01:09:54

Diagnostics Check		OK	01:00	
Air Blank		0.000	01:01	
Control Test		0.081	01:01	
Air Blank		0.000	01:02	
Subject Sample	# 1		01:05	No Sample Provided
Breath Volume		0.000	01:05	
Air Blank		0.000	01:05	
Air Blank		0.000	01:07	
Subject Sample	# 2		01:08	Subject Test Refused
Breath Volume		0.000	01:08	
Air Blank		0.000	01:08	
Control Test		0.081	01:09	
Air Blank		0.000	01:09	
Diagnostics Check		OK	01:09	

Data Download Date/Time: 05/29/2018 11:15

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASTINGS,KENNETH,S
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: HASTINGS,KENNETH
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 04/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: JOHNSTON

DATE/TIME OF TEST: 04/19/2018 01:18:45

Diagnostics Check		OK	01:13	
Air Blank		0.000	01:14	
Control Test		0.081	01:14	
Air Blank		0.000	01:15	
Subject Sample	# 1		01:17	Subject Test Refused
Breath Volume		0.000	01:17	
Air Blank		0.000	01:17	
Control Test		0.080	01:18	
Air Blank		0.000	01:18	
Diagnostics Check		OK	01:18	

Data Download Date/Time: 05/29/2018 11:15

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 04/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: BEATTY

DATE/TIME OF TEST: 04/20/2018 00:43:47

Diagnostics Check		OK	00:30	
Air Blank		0.000	00:30	
Control Test		0.080	00:31	
Air Blank		0.000	00:31	
Subject Sample	# 1	0.259	00:33	
Breath Volume		1.585	00:33	
Air Blank		0.000	00:33	
Air Blank		0.000	00:35	
Subject Sample	# 2	0.222	00:38	Volume Not Met
Breath Volume		0.050	00:38	
Air Blank		0.000	00:39	
Air Blank		0.000	00:41	
Subject Sample	# 3		00:42	Subject Test Refused
Breath Volume		0.000	00:42	
Air Blank		0.000	00:42	
Control Test		0.079	00:43	
Air Blank		0.000	00:43	
Diagnostics Check		OK	00:43	

Data Download Date/Time: 05/29/2018 11:15

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 04/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: CATHI BOLTEADA

DATE/TIME OF TEST: 04/20/2018 22:56:43

Diagnostics Check	OK	22:54	
Air Blank	0.000	22:55	
Control Test	0.079	22:55	
Air Blank	0.000	22:56	
		22:56	RFI Detect
Air Blank	0.000	22:56	

Data Download Date/Time: 05/29/2018 11:15

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 04/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: CATHI BOLTEADA

DATE/TIME OF TEST: 04/20/2018 23:07:39

Diagnostics Check		OK	22:58	
Air Blank		0.000	22:59	
Control Test		0.080	22:59	
Air Blank		0.000	23:00	
Subject Sample	# 1	0.045	23:03	Volume Not Met
Breath Volume		0.625	23:03	
Air Blank		0.000	23:03	
Air Blank		0.000	23:05	
Subject Sample	# 2		23:06	Subject Test Refused
Breath Volume		0.000	23:06	
Air Blank		0.000	23:06	
Control Test		0.080	23:07	
Air Blank		0.000	23:07	
Diagnostics Check		OK	23:07	

Data Download Date/Time: 05/29/2018 11:15

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASTINGS,KENNETH,S
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: HASTINGS,KENNETH
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 04/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: EGELHOFF

DATE/TIME OF TEST: 04/21/2018 23:09:42

Diagnostics Check		OK	22:59
Air Blank		0.000	23:00
Control Test		0.080	23:00
Air Blank		0.000	23:01
Subject Sample	# 1	0.079	23:03
Breath Volume		1.289	23:03
Air Blank		0.000	23:04
Air Blank		0.000	23:06
Subject Sample	# 2	0.079	23:08
Breath Volume		1.292	23:08
Air Blank		0.000	23:08
Control Test		0.080	23:09
Air Blank		0.000	23:09
Diagnostics Check		OK	23:09

Data Download Date/Time: 05/29/2018 11:15

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 04/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 04/27/2018 01:47:55

Diagnostics Check		OK	01:44	
Air Blank		0.000	01:45	
Control Test		0.080	01:45	
Air Blank		0.000	01:46	
Subject Sample	# 1		01:46	Subject Test Refused
Breath Volume		0.000	01:46	
Air Blank		0.000	01:47	
Control Test		0.081	01:47	
Air Blank		0.000	01:47	
Diagnostics Check		OK	01:47	

Data Download Date/Time: 05/29/2018 11:15

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 04/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: SERRANO RUIZ

DATE/TIME OF TEST: 04/28/2018 03:29:20

Diagnostics Check		OK	03:19
Air Blank		0.000	03:20
Control Test		0.081	03:20
Air Blank		0.000	03:21
Subject Sample	# 1	0.141	03:24
Breath Volume		1.179	03:24
Air Blank		0.000	03:24
Air Blank		0.000	03:26
Subject Sample	# 2	0.139	03:27
Breath Volume		2.261	03:27
Air Blank		0.000	03:28
Control Test		0.079	03:28
Air Blank		0.000	03:29
Diagnostics Check		OK	03:29

Data Download Date/Time: 05/29/2018 11:15

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASTINGS,KENNETH,S
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: HASTINGS,KENNETH
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 04/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:37

SUBJECT'S LAST NAME: MATCHETTE

DATE/TIME OF TEST: 04/28/2018 22:04:56

Diagnostics Check		OK	21:58
Air Blank		0.000	21:59
Control Test		0.080	21:59
Air Blank		0.000	21:59
Subject Sample	# 1	0.148	22:00
Breath Volume		1.574	22:00
Air Blank		0.000	22:00
Air Blank		0.000	22:02
Subject Sample	# 2	0.144	22:03
Breath Volume		1.578	22:03
Air Blank		0.000	22:04
Control Test		0.080	22:04
Air Blank		0.000	22:04
Diagnostics Check		OK	22:04

Data Download Date/Time: 05/29/2018 11:15

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 04/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: PERKINS

DATE/TIME OF TEST: 04/29/2018 01:25:27

Diagnostics Check		OK	01:17
Air Blank		0.000	01:17
Control Test		0.080	01:17
Air Blank		0.000	01:18
Subject Sample	# 1	0.111	01:19
Breath Volume		1.417	01:19
Air Blank		0.000	01:20
Air Blank		0.000	01:22
Subject Sample	# 2	0.110	01:23
Breath Volume		1.671	01:23
Air Blank		0.000	01:24
Control Test		0.080	01:24
Air Blank		0.000	01:25
Diagnostics Check		OK	01:25

Data Download Date/Time: 05/29/2018 11:15

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILLIPS,JORDAN,W
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WOFFORD,CLINT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: COLORADO PALACIOS

DATE/TIME OF TEST: 04/01/2018 01:04:24

Diagnostics Check		OK	00:57
Air Blank		0.000	00:58
Control Test		0.079	00:58
Air Blank		0.000	00:59
Subject Sample	# 1	0.153	00:59
Breath Volume		3.496	00:59
Air Blank		0.000	01:00
Air Blank		0.000	01:02
Subject Sample	# 2	0.144	01:02
Breath Volume		2.445	01:02
Air Blank		0.000	01:03
Control Test		0.078	01:03
Air Blank		0.000	01:04
Diagnostics Check		OK	01:04

Data Download Date/Time: 04/30/2018 11:19

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JEUDY,RINALDO,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: EARLS,CHRISTIAN
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 03/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: REYES JR

DATE/TIME OF TEST: 04/01/2018 02:31:13

Diagnostics Check		OK	02:24
Air Blank		0.000	02:25
Control Test		0.079	02:25
Air Blank		0.000	02:26
Subject Sample	# 1	0.145	02:26
Breath Volume		2.464	02:26
Air Blank		0.000	02:27
Air Blank		0.000	02:29
Subject Sample	# 2	0.145	02:29
Breath Volume		2.296	02:29
Air Blank		0.000	02:30
Control Test		0.079	02:30
Air Blank		0.000	02:31
Diagnostics Check		OK	02:31

Data Download Date/Time: 04/30/2018 11:19

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHRISTELLO,CLIFFORD,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WOFFORD,CLINT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: COLORADO

DATE/TIME OF TEST: 04/01/2018 03:07:48

Diagnostics Check		OK	03:01
Air Blank		0.000	03:01
Control Test		0.079	03:02
Air Blank		0.000	03:02
Subject Sample	# 1	0.163	03:03
Breath Volume		3.050	03:03
Air Blank		0.000	03:04
Air Blank		0.000	03:05
Subject Sample	# 2	0.154	03:06
Breath Volume		3.031	03:06
Air Blank		0.000	03:06
Control Test		0.078	03:07
Air Blank		0.000	03:07
Diagnostics Check		OK	03:07

Data Download Date/Time: 04/30/2018 11:19

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHRISTELLO,CLIFFORD,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: CHRISTELLO,CLIFFORD
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 04/01/2018 19:29:01

Diagnostics Check		OK	19:17	
Air Blank		0.000	19:17	
Control Test		0.078	19:18	
Air Blank		0.000	19:18	
Subject Sample	# 1	0.000	19:21	Volume Not Met
Breath Volume		0.019	19:21	
Air Blank		0.000	19:22	
Air Blank		0.000	19:24	
Subject Sample	# 2	0.175	19:27	Volume Not Met
Breath Volume		0.019	19:27	
Air Blank		0.000	19:28	
Control Test		0.078	19:28	
Air Blank		0.000	19:28	
Diagnostics Check		OK	19:29	

Data Download Date/Time: 04/30/2018 11:19

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GETSEE,JOY,S
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WOFFORD,CLINT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 04/02/2018 03:47:21

Diagnostics Check		OK	03:40
Air Blank		0.000	03:40
Control Test		0.078	03:41
Air Blank		0.000	03:41
Subject Sample	# 1	0.130	03:42
Breath Volume		1.191	03:42
Air Blank		0.000	03:42
Air Blank		0.000	03:44
Subject Sample	# 2	0.124	03:45
Breath Volume		1.636	03:45
Air Blank		0.000	03:46
Control Test		0.079	03:46
Air Blank		0.000	03:47
Diagnostics Check		OK	03:47

Data Download Date/Time: 04/30/2018 11:19

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHRISTELLO,CLIFFORD,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: CHRISTELLO,CLIFFORD
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 04/06/2018 01:30:52

Diagnostics Check		OK	01:19	
Air Blank		0.000	01:19	
Control Test		0.078	01:20	
Air Blank		0.000	01:20	
Subject Sample	# 1		01:23	No Sample Provided
Breath Volume		0.000	01:23	
Air Blank		0.000	01:24	
Air Blank		0.000	01:26	
Subject Sample	# 2		01:29	No Sample Provided
Breath Volume		0.000	01:29	
Air Blank		0.000	01:29	
Control Test		0.079	01:30	
Air Blank		0.000	01:30	
Diagnostics Check		OK	01:30	

Data Download Date/Time: 04/30/2018 11:19

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUABA,JASON,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GUABA,JASON
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: ZUNIGA CRUZ

DATE/TIME OF TEST: 04/08/2018 01:19:59

Diagnostics Check		OK	01:13
Air Blank		0.000	01:14
Control Test		0.079	01:14
Air Blank		0.000	01:14
Subject Sample	# 1	0.133	01:15
Breath Volume		1.539	01:15
Air Blank		0.000	01:15
Air Blank		0.000	01:17
Subject Sample	# 2	0.129	01:18
Breath Volume		2.304	01:18
Air Blank		0.000	01:19
Control Test		0.077	01:19
Air Blank		0.000	01:19
Diagnostics Check		OK	01:19

Data Download Date/Time: 04/30/2018 11:19

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWES,ERIC,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LAWES,ERIC
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: NGUYEN

DATE/TIME OF TEST: 04/08/2018 05:01:44

Diagnostics Check		OK	04:47	
Air Blank		0.000	04:47	
Control Test		0.079	04:48	
Air Blank		0.000	04:48	
Subject Sample	# 1	0.074	04:51	Volume Not Met
Breath Volume		0.312	04:51	
Air Blank		0.000	04:52	
Air Blank		0.000	04:54	
Subject Sample	# 2	0.068	04:57	
Breath Volume		1.484	04:57	
Air Blank		0.000	04:57	
Air Blank		0.000	04:59	
Subject Sample	# 3	0.081	05:00	
Breath Volume		1.742	05:00	
Air Blank		0.000	05:00	
Control Test		0.078	05:01	
Air Blank		0.000	05:01	
Diagnostics Check		OK	05:01	

Data Download Date/Time: 04/30/2018 11:19

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWES,ERIC,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LAWES,ERIC
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: NGUYEN

DATE/TIME OF TEST: 04/08/2018 05:08:11

Diagnostics Check		OK	05:03	
Air Blank		0.000	05:04	
Control Test		0.078	05:04	
Air Blank		0.000	05:04	
Subject Sample	# 1		05:06	Subject Test Refused
Breath Volume		0.066	05:06	
Air Blank		0.000	05:07	
Control Test		0.079	05:07	
Air Blank		0.000	05:08	
Diagnostics Check		OK	05:08	

Data Download Date/Time: 04/30/2018 11:19

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOLLEY,KATRYNA,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SOLLEY,KATRYNA
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: FOWLER

DATE/TIME OF TEST: 04/11/2018 00:43:08

Diagnostics Check		OK	00:36
Air Blank		0.000	00:36
Control Test		0.078	00:37
Air Blank		0.000	00:37
Subject Sample	# 1	0.093	00:38
Breath Volume		4.027	00:38
Air Blank		0.000	00:39
Air Blank		0.000	00:40
Subject Sample	# 2	0.093	00:41
Breath Volume		3.640	00:41
Air Blank		0.000	00:42
Control Test		0.078	00:42
Air Blank		0.000	00:43
Diagnostics Check		OK	00:43

Data Download Date/Time: 04/30/2018 11:19

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWES,ERIC,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LAWES,ERIC
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 04/11/2018 23:06:07

Diagnostics Check		OK	22:59
Air Blank		0.000	22:59
Control Test		0.079	23:00
Air Blank		0.000	23:00
Subject Sample	# 1	0.192	23:01
Breath Volume		2.042	23:01
Air Blank		0.000	23:02
Air Blank		0.000	23:04
Subject Sample	# 2	0.193	23:04
Breath Volume		2.082	23:04
Air Blank		0.000	23:05
Control Test		0.078	23:05
Air Blank		0.000	23:06
Diagnostics Check		OK	23:06

Data Download Date/Time: 04/30/2018 11:19

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWES,ERIC,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ROBERTS,DALE
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: SYKES

DATE/TIME OF TEST: 04/12/2018 02:46:39

Diagnostics Check		OK	02:42	
Air Blank		0.000	02:42	
Control Test		0.078	02:42	
Air Blank		0.000	02:43	
Subject Sample	# 1		02:45	Subject Test Refused
Breath Volume		0.000	02:45	
Air Blank		0.000	02:45	
Control Test		0.079	02:46	
Air Blank		0.000	02:46	
Diagnostics Check		OK	02:46	

Data Download Date/Time: 04/30/2018 11:19

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GETSEE,JOY,S
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WOFFORD,CLINT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: CARBAJAL

DATE/TIME OF TEST: 04/14/2018 04:01:44

Diagnostics Check		OK	03:58	
Air Blank		0.000	03:59	
Control Test		0.079	03:59	
Air Blank		0.000	04:00	
Subject Sample	# 1		04:00	Subject Test Refused
Breath Volume		0.000	04:00	
Air Blank		0.000	04:00	
Control Test		0.079	04:01	
Air Blank		0.000	04:01	
Diagnostics Check		OK	04:01	

Data Download Date/Time: 04/30/2018 11:20

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILLIPS,JORDAN,W
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WOFFORD,CLINT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 06:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:35

SUBJECT'S LAST NAME: VILLENA

DATE/TIME OF TEST: 04/14/2018 09:08:02

Diagnostics Check		OK	09:01
Air Blank		0.000	09:02
Control Test		0.078	09:02
Air Blank		0.000	09:03
Subject Sample	# 1	0.121	09:03
Breath Volume		3.535	09:03
Air Blank		0.000	09:04
Air Blank		0.000	09:05
Subject Sample	# 2	0.124	09:06
Breath Volume		3.355	09:06
Air Blank		0.000	09:07
Control Test		0.079	09:07
Air Blank		0.000	09:07
Diagnostics Check		OK	09:08

Data Download Date/Time: 04/30/2018 11:20

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOONE,BRANDON,S
BREATH TEST OPERATOR AGENCY: Winter Haven PD
ARRESTING OFFICER: WIERZBA,NATHAN
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 03/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 04/14/2018 23:23:52

Diagnostics Check		OK	23:20	
Air Blank		0.000	23:20	
Control Test		0.078	23:20	
Air Blank		0.000	23:21	
Subject Sample	# 1		23:22	Subject Test Refused
Breath Volume		0.089	23:22	
Air Blank		0.000	23:22	
Control Test		0.079	23:23	
Air Blank		0.000	23:23	
Diagnostics Check		OK	23:23	

Data Download Date/Time: 04/30/2018 11:20

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES,STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SYKES,STEVEN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:41

SUBJECT'S LAST NAME: MORAZ

DATE/TIME OF TEST: 04/15/2018 20:12:32

Diagnostics Check		OK	20:03
Air Blank		0.000	20:04
Control Test		0.078	20:04
Air Blank		0.000	20:05
Subject Sample	# 1	0.282	20:06
Breath Volume		1.835	20:06
Air Blank		0.000	20:07
Air Blank		0.000	20:09
Subject Sample	# 2	0.287	20:10
Breath Volume		2.386	20:10
Air Blank		0.000	20:11
Control Test		0.077	20:11
Air Blank		0.000	20:12
Diagnostics Check		OK	20:12

Data Download Date/Time: 04/30/2018 11:20

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JEUDY,RINALDO,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WOFFORD,CLINT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: BEEBE

DATE/TIME OF TEST: 04/15/2018 23:05:42

Diagnostics Check		OK	23:02	
Air Blank		0.000	23:03	
Control Test		0.078	23:03	
Air Blank		0.000	23:03	
Subject Sample	# 1		23:04	Subject Test Refused
Breath Volume		0.000	23:04	
Air Blank		0.000	23:04	
Control Test		0.078	23:05	
Air Blank		0.000	23:05	
Diagnostics Check		OK	23:05	

Data Download Date/Time: 04/30/2018 11:20

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOLLEY,KATRYNA,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SOLLEY,KATRYNA
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: JOHNSTON

DATE/TIME OF TEST: 04/16/2018 03:33:58

Diagnostics Check		OK	03:31	
Air Blank		0.000	03:31	
Control Test		0.078	03:31	
Air Blank		0.000	03:32	
Subject Sample	# 1		03:32	Subject Test Refused
Breath Volume		0.000	03:32	
Air Blank		0.000	03:33	
Control Test		0.078	03:33	
Air Blank		0.000	03:33	
Diagnostics Check		OK	03:33	

Data Download Date/Time: 04/30/2018 11:20

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SINCLAIR,CAMERON,W
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SINCLAIR,CAMERON
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 16:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:53

SUBJECT'S LAST NAME: RATLIFF JR

DATE/TIME OF TEST: 04/16/2018 17:23:07

Diagnostics Check	OK	17:22	
Air Blank		17:22	RFI Detect
Air Blank	0.000	17:23	

Data Download Date/Time: 04/30/2018 11:20

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SINCLAIR,CAMEREON,W
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SINCLAIR,CAMERON
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 16:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:53

SUBJECT'S LAST NAME: RATLIFF JR.

DATE/TIME OF TEST: 04/16/2018 17:37:09

Diagnostics Check		OK	17:25	
Air Blank		0.000	17:26	
Control Test		0.078	17:26	
Air Blank		0.000	17:27	
Subject Sample	# 1		17:30	No Sample Provided
Breath Volume		0.000	17:30	
Air Blank		0.000	17:30	
Air Blank		0.000	17:32	
Subject Sample	# 2		17:35	No Sample Provided
Breath Volume		0.000	17:35	
Air Blank		0.000	17:36	
Control Test		0.078	17:36	
Air Blank		0.000	17:37	
Diagnostics Check		OK	17:37	

Data Download Date/Time: 04/30/2018 11:20

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWES,ERIC,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LAWES,ERIC
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 04/21/2018 01:09:28

Diagnostics Check		OK	01:02
Air Blank		0.000	01:02
Control Test		0.078	01:03
Air Blank		0.000	01:03
Subject Sample	# 1	0.119	01:04
Breath Volume		2.242	01:04
Air Blank		0.000	01:05
Air Blank		0.000	01:07
Subject Sample	# 2	0.117	01:07
Breath Volume		2.015	01:07
Air Blank		0.000	01:08
Control Test		0.078	01:08
Air Blank		0.000	01:09
Diagnostics Check		OK	01:09

Data Download Date/Time: 04/30/2018 11:20

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER,ADAM,H
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: POTTER,ADAM
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: FILIPPINI

DATE/TIME OF TEST: 04/21/2018 01:35:20

Diagnostics Check		OK	01:29
Air Blank		0.000	01:29
Control Test		0.079	01:29
Air Blank		0.000	01:30
Subject Sample	# 1	0.171	01:30
Breath Volume		1.578	01:30
Air Blank		0.000	01:31
Air Blank		0.000	01:33
Subject Sample	# 2	0.165	01:33
Breath Volume		2.363	01:33
Air Blank		0.000	01:34
Control Test		0.078	01:34
Air Blank		0.000	01:35
Diagnostics Check		OK	01:35

Data Download Date/Time: 04/30/2018 11:20

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SINCLAIR,CAMERON,W
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SINCLAIR,CAMERON
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: PETERSEN

DATE/TIME OF TEST: 04/21/2018 02:32:27

Diagnostics Check		OK	02:25
Air Blank		0.000	02:26
Control Test		0.078	02:26
Air Blank		0.000	02:27
Subject Sample	# 1	0.177	02:27
Breath Volume		2.921	02:27
Air Blank		0.000	02:28
Air Blank		0.000	02:30
Subject Sample	# 2	0.184	02:30
Breath Volume		2.925	02:30
Air Blank		0.000	02:31
Control Test		0.077	02:31
Air Blank		0.000	02:32
Diagnostics Check		OK	02:32

Data Download Date/Time: 04/30/2018 11:20

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COSTALES,ALEXANDRO,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: COSTALES,ALEXANDRO
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: PETERS

DATE/TIME OF TEST: 04/22/2018 02:08:06

Diagnostics Check		OK	01:56	
Air Blank		0.000	01:57	
Control Test		0.080	01:57	
Air Blank		0.000	01:57	
Subject Sample	# 1		02:01	No Sample Provided
Breath Volume		0.000	02:01	
Air Blank		0.000	02:01	
Air Blank		0.000	02:03	
Subject Sample	# 2		02:06	No Sample Provided
Breath Volume		0.000	02:06	
Air Blank		0.000	02:07	
Control Test		0.079	02:07	
Air Blank		0.000	02:08	
Diagnostics Check		OK	02:08	

Data Download Date/Time: 04/30/2018 11:20

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWES,ERIC,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LAWES,ERIC
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: CONNELL

DATE/TIME OF TEST: 04/22/2018 03:36:43

Diagnostics Check		OK	03:29
Air Blank		0.000	03:29
Control Test		0.079	03:30
Air Blank		0.000	03:30
Subject Sample	# 1	0.192	03:31
Breath Volume		1.539	03:31
Air Blank		0.000	03:32
Air Blank		0.000	03:34
Subject Sample	# 2	0.190	03:35
Breath Volume		1.367	03:35
Air Blank		0.000	03:35
Control Test		0.078	03:36
Air Blank		0.000	03:36
Diagnostics Check		OK	03:36

Data Download Date/Time: 04/30/2018 11:20

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES,STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SYKES,STEVEN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: GOULD

DATE/TIME OF TEST: 04/23/2018 23:27:22

Diagnostics Check		OK	23:19
Air Blank		0.000	23:20
Control Test		0.078	23:20
Air Blank		0.000	23:21
Subject Sample	# 1	0.184	23:22
Breath Volume		1.835	23:22
Air Blank		0.000	23:23
Air Blank		0.000	23:25
Subject Sample	# 2	0.186	23:25
Breath Volume		1.777	23:25
Air Blank		0.000	23:26
Control Test		0.077	23:26
Air Blank		0.000	23:27
Diagnostics Check		OK	23:27

Data Download Date/Time: 04/30/2018 11:20

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWES,ERIC,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LAWES,ERIC
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: ZACKOWSKI

DATE/TIME OF TEST: 04/25/2018 22:04:43

Diagnostics Check		OK	21:58
Air Blank		0.000	21:58
Control Test		0.078	21:58
Air Blank		0.000	21:59
Subject Sample	# 1	0.232	22:00
Breath Volume		2.453	22:00
Air Blank		0.000	22:00
Air Blank		0.000	22:02
Subject Sample	# 2	0.225	22:03
Breath Volume		1.613	22:03
Air Blank		0.000	22:03
Control Test		0.077	22:04
Air Blank		0.000	22:04
Diagnostics Check		OK	22:04

Data Download Date/Time: 04/30/2018 11:20

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER,ADAM,H
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: POTTER,ADAM
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: MOLINA

DATE/TIME OF TEST: 04/27/2018 01:46:46

Diagnostics Check		OK	01:36
Air Blank		0.000	01:37
Control Test		0.078	01:37
Air Blank		0.000	01:37
Subject Sample	# 1	0.214	01:41
Breath Volume		2.394	01:41
Air Blank		0.000	01:41
Air Blank		0.000	01:43
Subject Sample	# 2	0.207	01:45
Breath Volume		1.777	01:45
Air Blank		0.000	01:45
Control Test		0.076	01:46
Air Blank		0.000	01:46
Diagnostics Check		OK	01:46

Data Download Date/Time: 04/30/2018 11:20

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GETSEE,JOY,S
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WOFFORD,CLINT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:18

SUBJECT'S LAST NAME: HESTER

DATE/TIME OF TEST: 04/27/2018 23:40:01

Diagnostics Check		OK	23:37	
Air Blank		0.000	23:37	
Control Test		0.078	23:37	
Air Blank		0.000	23:38	
Subject Sample	# 1		23:38	Subject Test Refused
Breath Volume		0.000	23:38	
Air Blank		0.000	23:39	
Control Test		0.078	23:39	
Air Blank		0.000	23:39	
Diagnostics Check		OK	23:40	

Data Download Date/Time: 04/30/2018 11:20

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOONE,BRANDON,S
BREATH TEST OPERATOR AGENCY: Winter Haven PD
ARRESTING OFFICER: WOFFORD,CLINT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:52

SUBJECT'S LAST NAME: CHOQUETTE

DATE/TIME OF TEST: 04/28/2018 05:22:50

Diagnostics Check		OK	05:15
Air Blank		0.000	05:16
Control Test		0.078	05:16
Air Blank		0.000	05:17
Subject Sample	# 1	0.119	05:18
Breath Volume		2.839	05:18
Air Blank		0.000	05:18
Air Blank		0.000	05:20
Subject Sample	# 2	0.121	05:21
Breath Volume		1.726	05:21
Air Blank		0.000	05:21
Control Test		0.079	05:22
Air Blank		0.000	05:22
Diagnostics Check		OK	05:22

Data Download Date/Time: 04/30/2018 11:20

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUCAS,NATHANIEL,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LUCAS,NATHANIEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 04/29/2018 01:28:56

Diagnostics Check		OK	01:17	
Air Blank		0.000	01:18	
Control Test		0.079	01:18	
Air Blank		0.000	01:18	
Subject Sample	# 1		01:21	No Sample Provided
Breath Volume		0.000	01:21	
Air Blank		0.000	01:22	
Air Blank		0.000	01:24	
Subject Sample	# 2		01:27	No Sample Provided
Breath Volume		0.000	01:27	
Air Blank		0.000	01:28	
Control Test		0.079	01:28	
Air Blank		0.000	01:28	
Diagnostics Check		OK	01:28	

Data Download Date/Time: 04/30/2018 11:20

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUCAS,NATHANIEL,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WOFFORD,CLINT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: BABCOCK

DATE/TIME OF TEST: 04/29/2018 03:18:04

Diagnostics Check		OK	03:09
Air Blank		0.000	03:10
Control Test		0.079	03:10
Air Blank		0.000	03:11
Subject Sample	# 1	0.186	03:13
Breath Volume		1.238	03:13
Air Blank		0.000	03:13
Air Blank		0.000	03:15
Subject Sample	# 2	0.195	03:16
Breath Volume		1.398	03:16
Air Blank		0.000	03:17
Control Test		0.076	03:17
Air Blank		0.000	03:17
Diagnostics Check		OK	03:18

Data Download Date/Time: 04/30/2018 11:20

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHRISTELLO,CLIFFORD,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: CHRISTELLO,CLIFFORD
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: HARRELL

DATE/TIME OF TEST: 04/29/2018 04:55:21

Diagnostics Check		OK	04:49
Air Blank		0.000	04:49
Control Test		0.078	04:49
Air Blank		0.000	04:50
Subject Sample	# 1	0.169	04:50
Breath Volume		2.101	04:50
Air Blank		0.000	04:51
Air Blank		0.000	04:53
Subject Sample	# 2	0.159	04:53
Breath Volume		2.218	04:53
Air Blank		0.000	04:54
Control Test		0.078	04:54
Air Blank		0.000	04:55
Diagnostics Check		OK	04:55

Data Download Date/Time: 04/30/2018 11:20

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUCAS,NATHANIEL,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WOFFORD,CLINT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 04/29/2018 05:06:36

Diagnostics Check		OK	05:00
Air Blank		0.000	05:00
Control Test		0.077	05:01
Air Blank		0.000	05:01
Subject Sample	# 1	0.134	05:02
Breath Volume		1.929	05:02
Air Blank		0.000	05:02
Air Blank		0.000	05:04
Subject Sample	# 2	0.133	05:05
Breath Volume		1.667	05:05
Air Blank		0.000	05:05
Control Test		0.079	05:06
Air Blank		0.000	05:06
Diagnostics Check		OK	05:06

Data Download Date/Time: 04/30/2018 11:20

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JEUDY,RINALDO,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WOFFORD,CLINT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: ALEXIS

DATE/TIME OF TEST: 04/30/2018 03:33:31

Diagnostics Check		OK	03:19	
Air Blank		0.000	03:19	
Control Test		0.079	03:19	
Air Blank		0.000	03:20	
Subject Sample	# 1	0.165	03:23	Volume Not Met
Breath Volume		0.441	03:23	
Air Blank		0.000	03:24	
Air Blank		0.000	03:26	
Subject Sample	# 2	0.184	03:27	
Breath Volume		1.359	03:27	
Air Blank		0.000	03:28	
Air Blank		0.000	03:30	
Subject Sample	# 3	0.176	03:31	
Breath Volume		1.421	03:31	
Air Blank		0.000	03:32	
Control Test		0.076	03:32	
Air Blank		0.000	03:33	
Diagnostics Check		OK	03:33	

Data Download Date/Time: 04/30/2018 11:20

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CINELLI,ADAM,D
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: CINELLI,ADAM
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 03/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: KARAS

DATE/TIME OF TEST: 04/06/2018 04:37:23

Diagnostics Check		OK	04:27
Air Blank		0.000	04:27
Control Test		0.080	04:28
Air Blank		0.000	04:28
Subject Sample	# 1	0.202	04:29
Breath Volume		1.351	04:29
Air Blank		0.000	04:30
Air Blank		0.000	04:31
Subject Sample	# 2	0.224	04:32
Breath Volume		1.566	04:32
Air Blank		0.000	04:33
Air Blank		0.000	04:35
Subject Sample	# 3	0.215	04:35
Breath Volume		1.531	04:35
Air Blank		0.000	04:36
Control Test		0.078	04:36
Air Blank		0.000	04:37
Diagnostics Check		OK	04:37

Data Download Date/Time: 04/25/2018 13:12

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARQUHARSON,KEVIN,D
BREATH TEST OPERATOR AGENCY: Plantation PD
ARRESTING OFFICER: JACKSON,KAROLINE
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 03/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A2
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:23

SUBJECT'S LAST NAME: MCKENZIE

DATE/TIME OF TEST: 04/08/2018 05:47:06

Diagnostics Check		OK	05:44	
Air Blank		0.000	05:44	
Control Test		0.077	05:45	
Air Blank		0.000	05:45	
Subject Sample	# 1		05:45	Subject Test Refused
Breath Volume		0.000	05:45	
Air Blank		0.000	05:46	
Control Test		0.076	05:46	
Air Blank		0.000	05:47	
Diagnostics Check		OK	05:47	

Data Download Date/Time: 04/19/2018 11:10

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GARCIA,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: ALBERTO MARTINEZ

DATE/TIME OF TEST: 04/01/2018 02:53:28

Diagnostics Check		OK	02:47
Air Blank		0.000	02:47
Control Test		0.081	02:48
Air Blank		0.000	02:48
Subject Sample	# 1	0.083	02:49
Breath Volume		2.320	02:49
Air Blank		0.000	02:49
Air Blank		0.000	02:51
Subject Sample	# 2	0.083	02:51
Breath Volume		2.214	02:51
Air Blank		0.000	02:52
Control Test		0.080	02:52
Air Blank		0.000	02:53
Diagnostics Check		OK	02:53

Data Download Date/Time: 04/02/2018 14:40

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 03/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 04/01/2018 03:24:37

Diagnostics Check		OK	03:18
Air Blank		0.000	03:19
Control Test		0.080	03:19
Air Blank		0.000	03:19
Subject Sample	# 1	0.142	03:20
Breath Volume		2.257	03:20
Air Blank		0.000	03:20
Air Blank		0.000	03:22
Subject Sample	# 2	0.140	03:23
Breath Volume		2.144	03:23
Air Blank		0.000	03:23
Control Test		0.079	03:24
Air Blank		0.000	03:24
Diagnostics Check		OK	03:24

Data Download Date/Time: 04/02/2018 14:40

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: POSADA,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 04/01/2018 04:45:18

Diagnostics Check		OK	04:42	
Air Blank		0.000	04:42	
Control Test		0.080	04:43	
Air Blank		0.000	04:43	
Subject Sample	# 1		04:43	Subject Test Refused
Breath Volume		0.000	04:43	
Air Blank		0.000	04:44	
Control Test		0.080	04:44	
Air Blank		0.000	04:45	
Diagnostics Check		OK	04:45	

Data Download Date/Time: 04/02/2018 14:40

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HOOVER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 04/02/2018 02:41:15

Diagnostics Check		OK	02:35
Air Blank		0.000	02:35
Control Test		0.080	02:35
Air Blank		0.000	02:36
Subject Sample	# 1	0.078	02:36
Breath Volume		1.250	02:36
Air Blank		0.000	02:37
Air Blank		0.000	02:39
Subject Sample	# 2	0.090	02:39
Breath Volume		1.835	02:39
Air Blank		0.000	02:40
Control Test		0.080	02:40
Air Blank		0.000	02:41
Diagnostics Check		OK	02:41

Data Download Date/Time: 04/02/2018 14:40

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 04/04/2018 00:18:02

Diagnostics Check		OK	00:15	
Air Blank		0.000	00:15	
Control Test		0.080	00:15	
Air Blank		0.000	00:16	
Subject Sample	# 1		00:16	Subject Test Refused
Breath Volume		0.000	00:16	
Air Blank		0.000	00:17	
Control Test		0.080	00:17	
Air Blank		0.000	00:17	
Diagnostics Check		OK	00:18	

Data Download Date/Time: 04/17/2018 12:39

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAWSON,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:39

SUBJECT'S LAST NAME: PARKINS

DATE/TIME OF TEST: 04/04/2018 06:03:34

Diagnostics Check		OK	06:00	
Air Blank		0.000	06:01	
Control Test		0.079	06:01	
Air Blank		0.000	06:02	
Subject Sample	# 1		06:02	Subject Test Refused
Breath Volume		0.000	06:02	
Air Blank		0.000	06:02	
Control Test		0.079	06:02	
Air Blank		0.000	06:03	
Diagnostics Check		OK	06:03	

Data Download Date/Time: 04/17/2018 12:39

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: COWHDRUY,P
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: HAMMOND

DATE/TIME OF TEST: 04/06/2018 00:34:51

Diagnostics Check		OK	00:31	
Air Blank		0.000	00:32	
Control Test		0.081	00:32	
Air Blank		0.000	00:33	
Subject Sample	# 1		00:33	Subject Test Refused
Breath Volume		0.000	00:33	
Air Blank		0.000	00:33	
Control Test		0.080	00:34	
Air Blank		0.000	00:34	
Diagnostics Check		OK	00:34	

Data Download Date/Time: 04/17/2018 12:39

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MILLER,C
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: PITTMAN

DATE/TIME OF TEST: 04/06/2018 02:43:21

Diagnostics Check		OK	02:37
Air Blank		0.000	02:37
Control Test		0.081	02:37
Air Blank		0.000	02:38
Subject Sample	# 1	0.093	02:38
Breath Volume		1.398	02:38
Air Blank		0.000	02:39
Air Blank		0.000	02:41
Subject Sample	# 2	0.088	02:41
Breath Volume		1.550	02:41
Air Blank		0.000	02:42
Control Test		0.080	02:42
Air Blank		0.000	02:43
Diagnostics Check		OK	02:43

Data Download Date/Time: 04/17/2018 12:39

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BROWN,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: PRYER

DATE/TIME OF TEST: 04/06/2018 03:59:01

Diagnostics Check		OK	03:52
Air Blank		0.000	03:53
Control Test		0.080	03:53
Air Blank		0.000	03:54
Subject Sample	# 1	0.207	03:54
Breath Volume		1.804	03:54
Air Blank		0.000	03:55
Air Blank		0.000	03:57
Subject Sample	# 2	0.204	03:57
Breath Volume		1.308	03:57
Air Blank		0.000	03:58
Control Test		0.079	03:58
Air Blank		0.000	03:58
Diagnostics Check		OK	03:59

Data Download Date/Time: 04/17/2018 12:39

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: UPTON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:38

SUBJECT'S LAST NAME: CANO RUIZ

DATE/TIME OF TEST: 04/06/2018 06:06:20

Diagnostics Check		OK	05:59
Air Blank		0.000	05:59
Control Test		0.080	05:59
Air Blank		0.000	06:00
Subject Sample	# 1	0.193	06:01
Breath Volume		1.175	06:01
Air Blank		0.000	06:01
Air Blank		0.000	06:03
Subject Sample	# 2	0.183	06:04
Breath Volume		1.351	06:04
Air Blank		0.000	06:05
Control Test		0.080	06:05
Air Blank		0.000	06:06
Diagnostics Check		OK	06:06

Data Download Date/Time: 04/17/2018 12:39

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRYANT,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: REUTER

DATE/TIME OF TEST: 04/08/2018 00:28:30

Diagnostics Check		OK	00:22
Air Blank		0.000	00:22
Control Test		0.080	00:23
Air Blank		0.000	00:23
Subject Sample	# 1	0.000	00:24
Breath Volume		2.488	00:24
Air Blank		0.000	00:24
Air Blank		0.000	00:26
Subject Sample	# 2	0.000	00:27
Breath Volume		2.968	00:27
Air Blank		0.000	00:27
Control Test		0.080	00:27
Air Blank		0.000	00:28
Diagnostics Check		OK	00:28

Data Download Date/Time: 04/17/2018 12:39

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: MANN

DATE/TIME OF TEST: 04/10/2018 01:01:03

Diagnostics Check		OK	00:55
Air Blank		0.000	00:55
Control Test		0.079	00:55
Air Blank		0.000	00:56
Subject Sample	# 1	0.177	00:56
Breath Volume		3.066	00:56
Air Blank		0.000	00:57
Air Blank		0.000	00:59
Subject Sample	# 2	0.182	00:59
Breath Volume		2.382	00:59
Air Blank		0.000	01:00
Control Test		0.080	01:00
Air Blank		0.000	01:00
Diagnostics Check		OK	01:01

Data Download Date/Time: 04/17/2018 12:39

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TURPIN,D
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:55

SUBJECT'S LAST NAME: PETERS

DATE/TIME OF TEST: 04/11/2018 06:21:34

Diagnostics Check		OK	06:15
Air Blank		0.000	06:15
Control Test		0.080	06:16
Air Blank		0.000	06:16
Subject Sample	# 1	0.100	06:17
Breath Volume		2.757	06:17
Air Blank		0.000	06:17
Air Blank		0.000	06:19
Subject Sample	# 2	0.103	06:20
Breath Volume		2.582	06:20
Air Blank		0.000	06:20
Control Test		0.080	06:20
Air Blank		0.000	06:21
Diagnostics Check		OK	06:21

Data Download Date/Time: 04/17/2018 12:39

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: TOPNICK

DATE/TIME OF TEST: 04/12/2018 02:16:12

Diagnostics Check		OK	02:10
Air Blank		0.000	02:10
Control Test		0.080	02:10
Air Blank		0.000	02:11
Subject Sample	# 1	0.178	02:11
Breath Volume		4.152	02:11
Air Blank		0.000	02:12
Air Blank		0.000	02:14
Subject Sample	# 2	0.176	02:14
Breath Volume		4.175	02:14
Air Blank		0.000	02:15
Control Test		0.079	02:15
Air Blank		0.000	02:16
Diagnostics Check		OK	02:16

Data Download Date/Time: 04/17/2018 12:39

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: HICKMAN

DATE/TIME OF TEST: 04/12/2018 04:06:44

Diagnostics Check		OK	04:03	
Air Blank		0.000	04:04	
Control Test		0.079	04:04	
Air Blank		0.000	04:05	
Subject Sample	# 1		04:05	Subject Test Refused
Breath Volume		0.000	04:05	
Air Blank		0.000	04:05	
Control Test		0.079	04:06	
Air Blank		0.000	04:06	
Diagnostics Check		OK	04:06	

Data Download Date/Time: 04/17/2018 12:39

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,M
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: RENFROE

DATE/TIME OF TEST: 04/13/2018 03:11:18

Diagnostics Check		OK	03:08	
Air Blank		0.000	03:08	
Control Test		0.080	03:09	
Air Blank		0.000	03:09	
Subject Sample	# 1		03:09	Subject Test Refused
Breath Volume		0.000	03:09	
Air Blank		0.000	03:10	
Control Test		0.080	03:10	
Air Blank		0.000	03:11	
Diagnostics Check		OK	03:11	

Data Download Date/Time: 04/17/2018 12:39

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WEBER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: FRAZIER

DATE/TIME OF TEST: 04/15/2018 00:16:51

Diagnostics Check		OK	00:10
Air Blank		0.000	00:10
Control Test		0.080	00:10
Air Blank		0.000	00:11
Subject Sample	# 1	0.124	00:12
Breath Volume		1.242	00:12
Air Blank		0.000	00:13
Air Blank		0.000	00:14
Subject Sample	# 2	0.128	00:15
Breath Volume		1.132	00:15
Air Blank		0.000	00:15
Control Test		0.079	00:16
Air Blank		0.000	00:16
Diagnostics Check		OK	00:16

Data Download Date/Time: 04/17/2018 12:39

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRYANT,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: ATWELL

DATE/TIME OF TEST: 04/15/2018 01:25:51

Diagnostics Check		OK	01:14	
Air Blank		0.000	01:14	
Control Test		0.079	01:14	
Air Blank		0.000	01:15	
Subject Sample	# 1	0.141	01:18	Volume Not Met
Breath Volume		0.898	01:18	
Air Blank		0.000	01:19	
Air Blank		0.000	01:21	
Subject Sample	# 2	0.139	01:24	Volume Not Met
Breath Volume		0.792	01:24	
Air Blank		0.000	01:24	
Control Test		0.079	01:25	
Air Blank		0.000	01:25	
Diagnostics Check		OK	01:25	

Data Download Date/Time: 04/17/2018 12:39

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TORRES,W
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: DAVID

DATE/TIME OF TEST: 04/15/2018 03:45:20

Diagnostics Check		OK	03:38
Air Blank		0.000	03:39
Control Test		0.079	03:39
Air Blank		0.000	03:40
Subject Sample	# 1	0.056	03:40
Breath Volume		1.171	03:40
Air Blank		0.000	03:41
Air Blank		0.000	03:42
Subject Sample	# 2	0.063	03:43
Breath Volume		3.269	03:43
Air Blank		0.000	03:44
Control Test		0.080	03:44
Air Blank		0.000	03:45
Diagnostics Check		OK	03:45

Data Download Date/Time: 04/17/2018 12:39

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNSON,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: GAYLORD

DATE/TIME OF TEST: 04/15/2018 05:30:03

Diagnostics Check		OK	05:27	
Air Blank		0.000	05:27	
Control Test		0.080	05:27	
Air Blank		0.000	05:28	
Subject Sample	# 1		05:28	Subject Test Refused
Breath Volume		0.000	05:28	
Air Blank		0.000	05:29	
Control Test		0.080	05:29	
Air Blank		0.000	05:29	
Diagnostics Check		OK	05:30	

Data Download Date/Time: 04/17/2018 12:39

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: FRAGOSO RIOS

DATE/TIME OF TEST: 04/16/2018 00:54:44

Diagnostics Check		OK	00:51	
Air Blank		0.000	00:52	
Control Test		0.080	00:52	
Air Blank		0.000	00:53	
Subject Sample	# 1		00:53	Subject Test Refused
Breath Volume		0.000	00:53	
Air Blank		0.000	00:53	
Control Test		0.081	00:54	
Air Blank		0.000	00:54	
Diagnostics Check		OK	00:54	

Data Download Date/Time: 04/17/2018 12:40

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CAMACHO,K
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 04/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 04/21/2018 04:46:29

Diagnostics Check		OK	04:40
Air Blank		0.000	04:40
Control Test		0.081	04:41
Air Blank		0.000	04:41
Subject Sample	# 1	0.207	04:41
Breath Volume		3.082	04:41
Air Blank		0.000	04:42
Air Blank		0.000	04:44
Subject Sample	# 2	0.214	04:44
Breath Volume		3.355	04:44
Air Blank		0.000	04:45
Control Test		0.079	04:45
Air Blank		0.000	04:46
Diagnostics Check		OK	04:46

Data Download Date/Time: 05/01/2018 14:49

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LANE,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: PODHANY

DATE/TIME OF TEST: 04/22/2018 03:51:49

Diagnostics Check		OK	03:45
Air Blank		0.000	03:46
Control Test		0.081	03:46
Air Blank		0.000	03:46
Subject Sample	# 1	0.031	03:47
Breath Volume		1.792	03:47
Air Blank		0.000	03:47
Air Blank		0.000	03:49
Subject Sample	# 2	0.031	03:50
Breath Volume		2.695	03:50
Air Blank		0.000	03:50
Control Test		0.080	03:51
Air Blank		0.000	03:51
Diagnostics Check		OK	03:51

Data Download Date/Time: 05/01/2018 14:49

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BROWN,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: GATHRIGHT

DATE/TIME OF TEST: 04/25/2018 00:00:48

Diagnostics Check		OK	23:57	
Air Blank		0.000	23:58	
Control Test		0.080	23:58	
Air Blank		0.000	23:59	
Subject Sample	# 1		23:59	Subject Test Refused
Breath Volume		0.000	23:59	
Air Blank		0.000	23:59	
Control Test		0.080	00:00	
Air Blank		0.000	00:00	
Diagnostics Check		OK	00:00	

Data Download Date/Time: 05/01/2018 14:49

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BAI,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: CHURCH

DATE/TIME OF TEST: 04/25/2018 03:57:16

Diagnostics Check		OK	03:51
Air Blank		0.000	03:51
Control Test		0.081	03:51
Air Blank		0.000	03:52
Subject Sample	# 1	0.230	03:52
Breath Volume		3.097	03:52
Air Blank		0.000	03:53
Air Blank		0.000	03:55
Subject Sample	# 2	0.234	03:55
Breath Volume		3.367	03:55
Air Blank		0.000	03:56
Control Test		0.080	03:56
Air Blank		0.000	03:57
Diagnostics Check		OK	03:57

Data Download Date/Time: 05/01/2018 14:49

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DUNLAP-PREVATT,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 04/26/2018 05:22:08

Diagnostics Check		OK	05:15
Air Blank		0.000	05:16
Control Test		0.080	05:16
Air Blank		0.000	05:16
Subject Sample	# 1	0.081	05:17
Breath Volume		2.675	05:17
Air Blank		0.000	05:18
Air Blank		0.000	05:20
Subject Sample	# 2	0.072	05:20
Breath Volume		1.593	05:20
Air Blank		0.000	05:21
Control Test		0.081	05:21
Air Blank		0.000	05:22
Diagnostics Check		OK	05:22

Data Download Date/Time: 05/01/2018 14:49

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MARRERO,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: WELLMAN

DATE/TIME OF TEST: 04/28/2018 00:15:24

Diagnostics Check		OK	00:12	
Air Blank		0.000	00:13	
Control Test		0.080	00:13	
Air Blank		0.000	00:13	
Subject Sample	# 1		00:14	Subject Test Refused
Breath Volume		0.000	00:14	
Air Blank		0.000	00:14	
Control Test		0.080	00:14	
Air Blank		0.000	00:15	
Diagnostics Check		OK	00:15	

Data Download Date/Time: 05/01/2018 14:49

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 04/29/2018 00:43:51

Diagnostics Check		OK	00:37
Air Blank		0.000	00:38
Control Test		0.079	00:38
Air Blank		0.000	00:38
Subject Sample	# 1	0.166	00:39
Breath Volume		2.945	00:39
Air Blank		0.000	00:39
Air Blank		0.000	00:41
Subject Sample	# 2	0.164	00:42
Breath Volume		2.925	00:42
Air Blank		0.000	00:42
Control Test		0.080	00:43
Air Blank		0.000	00:43
Diagnostics Check		OK	00:43

Data Download Date/Time: 05/01/2018 14:49

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HOOVER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: GAINEY

DATE/TIME OF TEST: 04/29/2018 03:08:16

Diagnostics Check		OK	03:02
Air Blank		0.000	03:02
Control Test		0.079	03:02
Air Blank		0.000	03:03
Subject Sample	# 1	0.112	03:03
Breath Volume		3.187	03:03
Air Blank		0.000	03:04
Air Blank		0.000	03:06
Subject Sample	# 2	0.112	03:06
Breath Volume		3.250	03:06
Air Blank		0.000	03:07
Control Test		0.079	03:07
Air Blank		0.000	03:08
Diagnostics Check		OK	03:08

Data Download Date/Time: 05/01/2018 14:49

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FILBERT,E
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 04/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: PRINCE

DATE/TIME OF TEST: 04/29/2018 05:56:48

Diagnostics Check		OK	05:50
Air Blank		0.000	05:50
Control Test		0.079	05:51
Air Blank		0.000	05:51
Subject Sample	# 1	0.073	05:52
Breath Volume		5.078	05:52
Air Blank		0.000	05:52
Air Blank		0.000	05:54
Subject Sample	# 2	0.075	05:55
Breath Volume		5.101	05:55
Air Blank		0.000	05:55
Control Test		0.080	05:56
Air Blank		0.000	05:56
Diagnostics Check		OK	05:56

Data Download Date/Time: 05/01/2018 14:49

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,M
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 04/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: YAZJI

DATE/TIME OF TEST: 04/30/2018 01:21:41

Diagnostics Check		OK	01:18	
Air Blank		0.000	01:19	
Control Test		0.080	01:19	
Air Blank		0.000	01:20	
Subject Sample	# 1		01:20	Subject Test Refused
Breath Volume		0.000	01:20	
Air Blank		0.000	01:20	
Control Test		0.081	01:21	
Air Blank		0.000	01:21	
Diagnostics Check		OK	01:21	

Data Download Date/Time: 05/01/2018 14:49

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GARCIA,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: CALSMER

DATE/TIME OF TEST: 04/30/2018 23:58:29

Diagnostics Check		OK	23:55	
Air Blank		0.000	23:56	
Control Test		0.080	23:56	
Air Blank		0.000	23:56	
Subject Sample	# 1		23:57	Subject Test Refused
Breath Volume		0.000	23:57	
Air Blank		0.000	23:57	
Control Test		0.080	23:57	
Air Blank		0.000	23:58	
Diagnostics Check		OK	23:58	

Data Download Date/Time: 05/01/2018 14:49

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOWALSKI,KEVIN,P
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: KOWALSKI,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 11:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:15

SUBJECT'S LAST NAME: DETTY

DATE/TIME OF TEST: 04/10/2018 12:41:28

Diagnostics Check		OK	12:35
Air Blank		0.000	12:35
Control Test		0.081	12:36
Air Blank		0.000	12:36
Subject Sample	# 1	0.082	12:37
Breath Volume		1.828	12:37
Air Blank		0.000	12:37
Air Blank		0.000	12:39
Subject Sample	# 2	0.079	12:40
Breath Volume		4.003	12:40
Air Blank		0.000	12:40
Control Test		0.082	12:40
Air Blank		0.000	12:41
Diagnostics Check		OK	12:41

Data Download Date/Time: 05/02/2018 22:56

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOWALSKI,KEVIN,P
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: KOWALSKI,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 18:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:08

SUBJECT'S LAST NAME: SEAMAN

DATE/TIME OF TEST: 04/11/2018 19:35:17

Diagnostics Check		OK	19:28
Air Blank		0.000	19:28
Control Test		0.082	19:28
Air Blank		0.000	19:29
Subject Sample	# 1	0.009	19:30
Breath Volume		1.332	19:30
Air Blank		0.000	19:30
Air Blank		0.000	19:32
Subject Sample	# 2	0.012	19:33
Breath Volume		2.015	19:33
Air Blank		0.000	19:34
Control Test		0.082	19:34
Air Blank		0.000	19:35
Diagnostics Check		OK	19:35

Data Download Date/Time: 05/02/2018 22:56

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRELL,CHAD,A
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: HARRELL,CHAD
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 18:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:39

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 04/23/2018 19:06:57

Diagnostics Check		OK	18:59
Air Blank		0.000	19:00
Control Test		0.082	19:00
Air Blank		0.000	19:01
Subject Sample	# 1	0.090	19:01
Breath Volume		1.273	19:01
Air Blank		0.000	19:02
Air Blank		0.000	19:04
Subject Sample	# 2	0.083	19:05
Breath Volume		1.644	19:05
Air Blank		0.000	19:06
Control Test		0.083	19:06
Air Blank		0.000	19:06
Diagnostics Check		OK	19:06

Data Download Date/Time: 05/02/2018 22:56

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOWALSKI,KEVIN,P
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: KOWALSKI,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 21:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:24

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 04/27/2018 21:53:16

Diagnostics Check		OK	21:44
Air Blank		0.000	21:44
Control Test		0.082	21:45
Air Blank		0.000	21:45
Subject Sample	# 1	0.124	21:46
Breath Volume		1.175	21:46
Air Blank		0.000	21:47
Air Blank		0.000	21:49
Subject Sample	# 2	0.125	21:51
Breath Volume		1.441	21:51
Air Blank		0.000	21:52
Control Test		0.083	21:52
Air Blank		0.000	21:53
Diagnostics Check		OK	21:53

Data Download Date/Time: 05/02/2018 22:56

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOPER,WILLIAM,J
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: STACY,MICHAEL
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 03/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG616503
DRY GAS EXPIRY DATE: 06/13/2018
VIOLATION CODE: DUI
ARREST TIME: 15:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: FOSTER

DATE/TIME OF TEST: 04/24/2018 17:11:17

Diagnostics Check		OK	17:08	
Air Blank		0.000	17:08	
Control Test		0.076	17:09	
Air Blank		0.000	17:09	
Subject Sample	# 1		17:09	Subject Test Refused
Breath Volume		0.000	17:09	
Air Blank		0.000	17:10	
Control Test		0.076	17:10	
Air Blank		0.000	17:11	
Diagnostics Check		OK	17:11	

Data Download Date/Time: 04/30/2018 06:50

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALTER,CRAIG,K
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: WELCH,K
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 03/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG616503
DRY GAS EXPIRY DATE: 06/13/2018
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: HULL

DATE/TIME OF TEST: 04/29/2018 02:56:12

Diagnostics Check		OK	02:53	
Air Blank		0.000	02:53	
Control Test		0.075	02:54	
Air Blank		0.000	02:54	
Subject Sample	# 1		02:54	Subject Test Refused
Breath Volume		0.000	02:54	
Air Blank		0.000	02:55	
Control Test		0.076	02:55	
Air Blank		0.000	02:56	
Diagnostics Check		OK	02:56	

Data Download Date/Time: 04/30/2018 06:50

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALTER,CRAIG,K
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: BROWN,TREVOR
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 03/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG616503
DRY GAS EXPIRY DATE: 06/13/2018
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: DANIELS

DATE/TIME OF TEST: 04/29/2018 04:20:25

Diagnostics Check		OK	04:12
Air Blank		0.000	04:13
Control Test		0.076	04:13
Air Blank		0.000	04:14
Subject Sample	# 1	0.098	04:15
Breath Volume		2.175	04:15
Air Blank		0.000	04:16
Air Blank		0.000	04:18
Subject Sample	# 2	0.095	04:18
Breath Volume		1.277	04:18
Air Blank		0.000	04:19
Control Test		0.075	04:19
Air Blank		0.000	04:20
Diagnostics Check		OK	04:20

Data Download Date/Time: 04/30/2018 06:50

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANE,MARC,T
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: LANE,MARC
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 04/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 04/26/2018 00:24:50

Diagnostics Check		OK	00:18
Air Blank		0.000	00:18
Control Test		0.079	00:19
Air Blank		0.000	00:19
Subject Sample	# 1	0.183	00:20
Breath Volume		2.664	00:20
Air Blank		0.000	00:20
Air Blank		0.000	00:22
Subject Sample	# 2	0.174	00:23
Breath Volume		2.894	00:23
Air Blank		0.000	00:23
Control Test		0.079	00:24
Air Blank		0.000	00:24
Diagnostics Check		OK	00:24

Data Download Date/Time: 05/29/2018 17:15

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANE,MARC,T
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: LANE,MARC
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 04/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: KANE

DATE/TIME OF TEST: 04/26/2018 23:17:12

Diagnostics Check		OK	23:11
Air Blank		0.000	23:11
Control Test		0.080	23:11
Air Blank		0.000	23:12
Subject Sample	# 1	0.227	23:12
Breath Volume		2.578	23:12
Air Blank		0.000	23:13
Air Blank		0.000	23:15
Subject Sample	# 2	0.210	23:15
Breath Volume		1.539	23:15
Air Blank		0.000	23:16
Control Test		0.079	23:16
Air Blank		0.000	23:17
Diagnostics Check		OK	23:17

Data Download Date/Time: 05/29/2018 17:15

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BECKMAN,KEVIN,E
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: GOODWALD,DARREN
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 04/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: FALLON

DATE/TIME OF TEST: 04/27/2018 23:47:12

Diagnostics Check		OK	23:40
Air Blank		0.000	23:40
Control Test		0.081	23:41
Air Blank		0.000	23:41
Subject Sample	# 1	0.133	23:42
Breath Volume		3.406	23:42
Air Blank		0.000	23:43
Air Blank		0.000	23:45
Subject Sample	# 2	0.122	23:45
Breath Volume		3.406	23:45
Air Blank		0.000	23:46
Control Test		0.079	23:46
Air Blank		0.000	23:47
Diagnostics Check		OK	23:47

Data Download Date/Time: 05/29/2018 17:15

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,WILLIAM,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: NIXON,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 14:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:05

SUBJECT'S LAST NAME: RANGAR

DATE/TIME OF TEST: 04/28/2018 15:33:41

Diagnostics Check		OK	15:25	
Air Blank		0.000	15:26	
Control Test		0.081	15:26	
Air Blank		0.000	15:27	
Subject Sample	# 1	0.213	15:27	
Breath Volume		1.519	15:27	
Air Blank		0.000	15:28	
Air Blank		0.000	15:29	
Subject Sample	# 2		15:32	Subject Test Refused
Breath Volume		0.968	15:32	
Air Blank		0.000	15:32	
Control Test		0.080	15:33	
Air Blank		0.000	15:33	
Diagnostics Check		OK	15:33	

Data Download Date/Time: 05/29/2018 17:15

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COFFEE,JASON,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: COFFEE,JASON
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: WARKOW

DATE/TIME OF TEST: 04/29/2018 23:33:53

Diagnostics Check		OK	23:22	
Air Blank		0.000	23:22	
Control Test		0.081	23:23	
Air Blank		0.000	23:23	
Subject Sample	# 1	0.051	23:26	Volume Not Met
Breath Volume		0.296	23:26	
Air Blank		0.000	23:27	
Air Blank		0.000	23:29	
Subject Sample	# 2	0.066	23:32	Volume Not Met
Breath Volume		0.097	23:32	
Air Blank		0.000	23:32	
Control Test		0.078	23:33	
Air Blank		0.000	23:33	
Diagnostics Check		OK	23:33	

Data Download Date/Time: 05/29/2018 17:15

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAEZ HERNANDEZ,DENYS,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: PAEZ HERNANDEZ,DENYS
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: NOONAN

DATE/TIME OF TEST: 04/30/2018 00:00:54

Diagnostics Check		OK	23:52	
Air Blank		0.000	23:52	
Control Test		0.080	23:53	
Air Blank		0.000	23:53	
Subject Sample	# 1		23:56	No Sample Provided
Breath Volume		0.000	23:56	
Air Blank		0.000	23:57	
Air Blank		0.000	23:59	
Subject Sample	# 2		23:59	Subject Test Refused
Breath Volume		0.000	23:59	
Air Blank		0.000	00:00	
Control Test		0.080	00:00	
Air Blank		0.000	00:00	
Diagnostics Check		OK	00:00	

Data Download Date/Time: 05/29/2018 17:15

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRIOLAT,CASEY,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BRIOLAT,CASEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:16

SUBJECT'S LAST NAME: AMANTE

DATE/TIME OF TEST: 04/30/2018 22:45:24

Diagnostics Check		OK	22:38
Air Blank		0.000	22:38
Control Test		0.080	22:39
Air Blank		0.000	22:39
Subject Sample	# 1	0.267	22:40
Breath Volume		2.304	22:40
Air Blank		0.000	22:41
Air Blank		0.000	22:43
Subject Sample	# 2	0.275	22:43
Breath Volume		2.097	22:43
Air Blank		0.000	22:44
Control Test		0.080	22:44
Air Blank		0.000	22:45
Diagnostics Check		OK	22:45

Data Download Date/Time: 05/29/2018 17:15

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARGE,SUSAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BARGE,SUSAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: BEACH

DATE/TIME OF TEST: 04/01/2018 03:19:38

Diagnostics Check		OK	03:13
Air Blank		0.000	03:13
Control Test		0.080	03:13
Air Blank		0.000	03:14
Subject Sample	# 1	0.097	03:14
Breath Volume		1.167	03:14
Air Blank		0.000	03:15
Air Blank		0.000	03:17
Subject Sample	# 2	0.110	03:18
Breath Volume		1.910	03:18
Air Blank		0.000	03:18
Control Test		0.082	03:19
Air Blank		0.000	03:19
Diagnostics Check		OK	03:19

Data Download Date/Time: 04/05/2018 20:27

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARGE,SUSAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BARGE,SUSAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: CHAPALA

DATE/TIME OF TEST: 04/06/2018 23:02:53

Diagnostics Check		OK	22:56
Air Blank		0.000	22:56
Control Test		0.080	22:57
Air Blank		0.000	22:57
Subject Sample	# 1	0.146	22:58
Breath Volume		5.257	22:58
Air Blank		0.000	22:58
Air Blank		0.000	23:00
Subject Sample	# 2	0.144	23:01
Breath Volume		5.285	23:01
Air Blank		0.000	23:01
Control Test		0.080	23:02
Air Blank		0.000	23:02
Diagnostics Check		OK	23:02

Data Download Date/Time: 05/08/2018 15:00

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARGE,SUSAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BARGE,SUSAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: PANGBURN

DATE/TIME OF TEST: 04/07/2018 23:44:53

Diagnostics Check		OK	23:37
Air Blank		0.000	23:38
Control Test		0.080	23:38
Air Blank		0.000	23:39
Subject Sample	# 1	0.170	23:40
Breath Volume		1.824	23:40
Air Blank		0.000	23:40
Air Blank		0.000	23:42
Subject Sample	# 2	0.160	23:43
Breath Volume		2.046	23:43
Air Blank		0.000	23:43
Control Test		0.081	23:44
Air Blank		0.000	23:44
Diagnostics Check		OK	23:44

Data Download Date/Time: 05/08/2018 15:00

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARGE,SUSAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BARGE,SUSAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: BELINGE

DATE/TIME OF TEST: 04/08/2018 04:04:29

Diagnostics Check		OK	03:48	
Air Blank		0.000	03:49	
Control Test		0.082	03:49	
Air Blank		0.000	03:49	
Subject Sample	# 1	0.170	03:51	
Breath Volume		2.074	03:51	
Air Blank		0.000	03:52	
Air Blank		0.000	03:54	
Subject Sample	# 2	0.129	03:57	Volume Not Met
Breath Volume		0.140	03:57	
Air Blank		0.000	03:57	
Air Blank		0.000	03:59	
Subject Sample	# 3	0.137	04:02	Volume Not Met
Breath Volume		0.500	04:02	
Air Blank		0.000	04:03	
Control Test		0.080	04:03	
Air Blank		0.000	04:04	
Diagnostics Check		OK	04:04	

Data Download Date/Time: 05/08/2018 15:00

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARGE,SUSAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BARGE,SUSAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: BELINGE

DATE/TIME OF TEST: 04/08/2018 04:19:13

Diagnostics Check		OK	04:06	
Air Blank		0.000	04:07	
Control Test		0.079	04:07	
Air Blank		0.000	04:07	
Subject Sample	# 1	0.161	04:09	No .020 agreement
Breath Volume		2.179	04:09	
Air Blank		0.000	04:09	
Air Blank		0.000	04:11	
Subject Sample	# 2	0.147	04:14	Volume Not Met
Breath Volume		0.000	04:14	
Air Blank		0.000	04:15	
Air Blank		0.000	04:17	
Subject Sample	# 3	0.137	04:17	No .020 agreement
Breath Volume		1.589	04:17	
Air Blank		0.000	04:18	
Control Test		0.078	04:18	
Air Blank		0.000	04:19	
Diagnostics Check		OK	04:19	

Data Download Date/Time: 05/08/2018 15:00

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARGE,SUSAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BARGE,SUSAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: BELINGE

DATE/TIME OF TEST: 04/08/2018 04:28:57

Diagnostics Check		OK	04:22
Air Blank		0.000	04:23
Control Test		0.078	04:23
Air Blank		0.000	04:24
Subject Sample	# 1	0.159	04:24
Breath Volume		1.343	04:24
Air Blank		0.000	04:24
Air Blank		0.000	04:26
Subject Sample	# 2	0.159	04:27
Breath Volume		1.597	04:27
Air Blank		0.000	04:28
Control Test		0.081	04:28
Air Blank		0.000	04:28
Diagnostics Check		OK	04:28

Data Download Date/Time: 05/08/2018 15:00

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARGE,SUSAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BARGE,SUSAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:59

SUBJECT'S LAST NAME: BANKS

DATE/TIME OF TEST: 04/15/2018 17:25:34

Diagnostics Check		OK	17:19	
Air Blank		0.000	17:19	
Control Test		0.079	17:20	
Air Blank		0.000	17:20	
Subject Sample	# 1		17:23	No Sample Provided
Breath Volume		0.000	17:23	
Air Blank			17:24	RFI Detect
Air Blank		0.000	17:24	
Control Test		0.079	17:24	
Air Blank		0.000	17:25	
Diagnostics Check		OK	17:25	

Data Download Date/Time: 05/08/2018 15:00

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARGE,SUSAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BARGE,SUSAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:59

SUBJECT'S LAST NAME: BANKS

DATE/TIME OF TEST: 04/15/2018 17:36:17

Diagnostics Check		OK	17:27	
Air Blank		0.000	17:28	
Control Test		0.081	17:28	
Air Blank		0.000	17:29	
Subject Sample	# 1		17:32	No Sample Provided
Breath Volume		0.000	17:32	
Air Blank		0.000	17:32	
Air Blank		0.000	17:34	
Subject Sample	# 2		17:34	Subject Test Refused
Breath Volume		0.000	17:34	
Air Blank		0.000	17:35	
Control Test		0.080	17:35	
Air Blank		0.000	17:36	
Diagnostics Check		OK	17:36	

Data Download Date/Time: 05/08/2018 15:00

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001128

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRENCH,CALEB,L
BREATH TEST OPERATOR AGENCY: Putnam County SO
ARRESTING OFFICER: BLOUNT,DONALD
ARRESTING OFFICER AGENCY: Putnam County SO
LAST AGENCY INSPECTION: 03/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A3
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 12:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: COSTON

DATE/TIME OF TEST: 04/01/2018 02:54:31

Diagnostics Check		OK	02:51	
Air Blank		0.000	02:52	
Control Test		0.080	02:52	
Air Blank		0.000	02:53	
Subject Sample	# 1		02:53	Subject Test Refused
Breath Volume		0.000	02:53	
Air Blank		0.000	02:53	
Control Test		0.079	02:53	
Air Blank		0.000	02:54	
Diagnostics Check		OK	02:54	

Data Download Date/Time: 04/02/2018 17:42

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COKER,BRYAN,C
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: MYERS,MATTHEW
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 03/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: SALVATORE

DATE/TIME OF TEST: 04/07/2018 02:50:14

Diagnostics Check		OK	02:43
Air Blank		0.000	02:44
Control Test		0.078	02:44
Air Blank		0.000	02:45
Subject Sample	# 1	0.131	02:45
Breath Volume		1.535	02:45
Air Blank		0.000	02:46
Air Blank		0.000	02:48
Subject Sample	# 2	0.130	02:48
Breath Volume		2.449	02:48
Air Blank		0.000	02:49
Control Test		0.077	02:49
Air Blank		0.000	02:50
Diagnostics Check		OK	02:50

Data Download Date/Time: 04/16/2018 12:20

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,ROBERT,J
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: ZIGLER,RICK
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 09:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:11

SUBJECT'S LAST NAME: PUTCHACONIS

DATE/TIME OF TEST: 04/28/2018 11:38:12

Diagnostics Check		OK	11:31
Air Blank		0.000	11:32
Control Test		0.079	11:32
Air Blank		0.000	11:33
Subject Sample	# 1	0.341	11:33
Breath Volume		2.167	11:33
Air Blank		0.000	11:34
Air Blank		0.000	11:36
Subject Sample	# 2	0.323	11:36
Breath Volume		2.976	11:36
Air Blank		0.000	11:37
Control Test		0.079	11:37
Air Blank		0.000	11:38
Diagnostics Check		OK	11:38

Data Download Date/Time: 05/30/2018 11:40

Printed: 12 July 2018

COBRA ©1997-2014 CMI, Inc.