

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVENDANO,JUAN,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: DEKEYSER,NIKOLAS
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: MEJIA

DATE/TIME OF TEST: 05/01/2018 02:37:47

Diagnostics Check	OK	02:30
Air Blank	0.000	02:30
Control Test	0.082	02:30
Air Blank	0.000	02:31
Subject Sample # 1	0.170	02:32
Breath Volume	2.601	02:32
Air Blank	0.000	02:33
Air Blank	0.000	02:34
Subject Sample # 2	0.177	02:36
Breath Volume	2.687	02:36
Air Blank	0.000	02:36
Control Test	0.080	02:37
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

Data Download Date/Time: 05/09/2018 12:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVENDANO,JUAN,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SELIMOVIC,ADNAN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: RIVERA DE LA CRUZ

DATE/TIME OF TEST: 05/02/2018 03:45:03

Diagnostics Check	OK	03:37
Air Blank	0.000	03:38
Control Test	0.082	03:38
Air Blank	0.000	03:38
Subject Sample # 1	0.232	03:40
Breath Volume	2.265	03:40
Air Blank	0.000	03:41
Air Blank	0.000	03:42
Subject Sample # 2	0.227	03:43
Breath Volume	1.828	03:43
Air Blank	0.000	03:44
Control Test	0.081	03:44
Air Blank	0.000	03:44
Diagnostics Check	OK	03:45

Data Download Date/Time: 05/09/2018 12:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,J,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: RIVERA,J
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: HERRERA-IXCOY

DATE/TIME OF TEST: 05/02/2018 23:18:04

Diagnostics Check	OK	23:11
Air Blank	0.000	23:12
Control Test	0.081	23:12
Air Blank	0.000	23:12
Subject Sample # 1	0.132	23:13
Breath Volume	1.871	23:13
Air Blank	0.000	23:14
Air Blank	0.000	23:15
Subject Sample # 2	0.133	23:16
Breath Volume	2.167	23:16
Air Blank	0.000	23:17
Control Test	0.081	23:17
Air Blank	0.000	23:17
Diagnostics Check	OK	23:18

Data Download Date/Time: 05/09/2018 12:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONSALL, MATTHEW, A
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: BELLINI, R
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 05/05/2018 04:41:33

Diagnostics Check	OK	04:34
Air Blank	0.000	04:35
Control Test	0.081	04:35
Air Blank	0.000	04:36
Subject Sample # 1	0.188	04:36
Breath Volume	3.472	04:36
Air Blank	0.000	04:37
Air Blank	0.000	04:39
Subject Sample # 2	0.180	04:39
Breath Volume	1.566	04:39
Air Blank	0.000	04:40
Control Test	0.080	04:40
Air Blank	0.000	04:41
Diagnostics Check	OK	04:41

Data Download Date/Time: 05/09/2018 12:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNCAN,N,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: DUNCAN,N
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:25

SUBJECT'S LAST NAME: CARLTON

DATE/TIME OF TEST: 05/05/2018 18:56:18

Diagnostics Check	OK	18:48
Air Blank	0.000	18:49
Control Test	0.081	18:49
Air Blank	0.000	18:50
Subject Sample # 1	0.187	18:51
Breath Volume	2.011	18:51
Air Blank	0.000	18:52
Air Blank	0.000	18:53
Subject Sample # 2	0.194	18:54
Breath Volume	1.765	18:54
Air Blank	0.000	18:55
Control Test	0.080	18:55
Air Blank	0.000	18:56
Diagnostics Check	OK	18:56

Data Download Date/Time: 05/09/2018 12:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSEPH,G
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: RYAN,MATTHEW
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 05/06/2018 03:52:25

Diagnostics Check	OK	03:44
Air Blank	0.000	03:44
Control Test	0.081	03:45
Air Blank	0.000	03:45
Subject Sample # 1	0.161	03:47
Breath Volume	1.417	03:47
Air Blank	0.000	03:48
Air Blank	0.000	03:49
Subject Sample # 2	0.167	03:50
Breath Volume	1.886	03:50
Air Blank	0.000	03:51
Control Test	0.079	03:51
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time: 05/09/2018 12:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUCHENEY, WESLEY, A
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: REQUENA, CARLOS
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 04/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: PLAYER

DATE/TIME OF TEST: 05/01/2018 02:50:11

Diagnostics Check	OK	02:42
Air Blank	0.000	02:42
Control Test	0.079	02:43
Air Blank	0.000	02:43
Subject Sample # 1	0.198	02:45
Breath Volume	1.625	02:45
Air Blank	0.000	02:46
Air Blank	0.000	02:48
Subject Sample # 2	0.202	02:48
Breath Volume	2.128	02:48
Air Blank	0.000	02:49
Control Test	0.077	02:49
Air Blank	0.000	02:50
Diagnostics Check	OK	02:50

Data Download Date/Time: 05/04/2018 10:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOLBICKA,LAWRENCE,R
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SCHIBLER,BRIAN
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 04/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 11:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:26

SUBJECT'S LAST NAME: FRAZIER

DATE/TIME OF TEST: 05/01/2018 12:53:04

Diagnostics Check	OK	12:47
Air Blank	0.000	12:47
Control Test	0.079	12:48
Air Blank	0.000	12:48
Subject Sample # 1	0.000	12:48
Breath Volume	2.324	12:48
Air Blank	0.000	12:49
Air Blank	0.000	12:51
Subject Sample # 2	0.000	12:51
Breath Volume	2.246	12:51
Air Blank	0.000	12:52
Control Test	0.079	12:52
Air Blank	0.000	12:52
Diagnostics Check	OK	12:53

Data Download Date/Time: 05/04/2018 10:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEA,JOHN,D
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SHEA,JOHN
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 04/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: HUYNH

DATE/TIME OF TEST: 05/01/2018 23:03:36

Diagnostics Check	OK	22:48	
Air Blank	0.000	22:48	
Control Test	0.078	22:48	
Air Blank	0.000	22:49	
Subject Sample	# 1 0.140	22:52	
Breath Volume	1.503	22:52	
Air Blank	0.000	22:53	
Air Blank	0.000	22:54	
Subject Sample	# 2	22:57	No Sample Provided
Breath Volume	0.000	22:57	
Air Blank	0.000	22:58	
Air Blank	0.000	23:00	
Subject Sample	# 3	23:02	Subject Test Refused
Breath Volume	0.000	23:02	
Air Blank	0.000	23:02	
Control Test	0.078	23:03	
Air Blank	0.000	23:03	
Diagnostics Check	OK	23:03	

Data Download Date/Time: 05/04/2018 10:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SKINNER,SEAN
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 04/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: RICHARDS

DATE/TIME OF TEST: 05/04/2018 02:51:41

Diagnostics Check	OK	02:48	
Air Blank	0.000	02:49	
Control Test	0.079	02:49	
Air Blank	0.000	02:50	
Subject Sample	# 1	02:50	Subject Test Refused
Breath Volume	0.000	02:50	
Air Blank	0.000	02:50	
Control Test	0.078	02:51	
Air Blank	0.000	02:51	
Diagnostics Check	OK	02:51	

Data Download Date/Time: 05/04/2018 10:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JASIUKIEWICZ,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 05/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: BALDIVIA

DATE/TIME OF TEST: 05/05/2018 00:10:49

Diagnostics Check	OK	00:04
Air Blank	0.000	00:05
Control Test	0.079	00:05
Air Blank	0.000	00:06
Subject Sample # 1	0.000	00:06
Breath Volume	1.617	00:06
Air Blank	0.000	00:07
Air Blank	0.000	00:09
Subject Sample # 2	0.000	00:09
Breath Volume	1.167	00:09
Air Blank	0.000	00:09
Control Test	0.078	00:10
Air Blank	0.000	00:10
Diagnostics Check	OK	00:10

Data Download Date/Time: 06/08/2018 13:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JASIUKIEWICZ,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 05/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:14

SUBJECT'S LAST NAME: STANLEY

DATE/TIME OF TEST: 05/05/2018 21:36:58

Diagnostics Check	OK	21:34	
Air Blank	0.000	21:34	
Control Test	0.079	21:34	
Air Blank	0.000	21:35	
Subject Sample	# 1	21:35	Subject Test Refused
Breath Volume	0.000	21:35	
Air Blank	0.000	21:36	
Control Test	0.078	21:36	
Air Blank	0.000	21:36	
Diagnostics Check	OK	21:36	

Data Download Date/Time: 06/08/2018 13:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOYLE,THOMAS,E
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: BOYLE,THOMAS
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 05/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: ALFINO

DATE/TIME OF TEST: 05/06/2018 02:34:16

Diagnostics Check	OK	02:25
Air Blank	0.000	02:25
Control Test	0.078	02:25
Air Blank	0.000	02:26
Subject Sample # 1	0.084	02:29
Breath Volume	1.585	02:29
Air Blank	0.000	02:30
Air Blank	0.000	02:32
Subject Sample # 2	0.084	02:32
Breath Volume	1.285	02:32
Air Blank	0.000	02:33
Control Test	0.076	02:33
Air Blank	0.000	02:34
Diagnostics Check	OK	02:34

Data Download Date/Time: 06/08/2018 13:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEA,JOHN,D
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SHEA,JOHN
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 05/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 05/06/2018 03:33:20

Diagnostics Check	OK	03:30	
Air Blank	0.000	03:30	
Control Test	0.078	03:31	
Air Blank	0.000	03:31	
Subject Sample	# 1	03:31	Subject Test Refused
Breath Volume	0.000	03:31	
Air Blank	0.000	03:32	
Control Test	0.077	03:32	
Air Blank	0.000	03:33	
Diagnostics Check	OK	03:33	

Data Download Date/Time: 06/08/2018 13:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEA,JOHN,D
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SHEA,JOHN
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 05/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:10

SUBJECT'S LAST NAME: VEGA

DATE/TIME OF TEST: 05/10/2018 17:36:56

Diagnostics Check	OK	17:31
Air Blank	0.000	17:31
Control Test	0.079	17:31
Air Blank	0.000	17:32
Subject Sample # 1	0.000	17:32
Breath Volume	3.421	17:32
Air Blank	0.000	17:33
Air Blank	0.000	17:35
Subject Sample # 2	0.000	17:35
Breath Volume	2.925	17:35
Air Blank	0.000	17:36
Control Test	0.078	17:36
Air Blank	0.000	17:36
Diagnostics Check	OK	17:36

Data Download Date/Time: 06/08/2018 13:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEAWARD,MORGANA,A
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: GUY,JOSHUA
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 05/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: MANZULLO

DATE/TIME OF TEST: 05/11/2018 01:02:17

Diagnostics Check	OK	00:59	
Air Blank	0.000	00:59	
Control Test	0.078	01:00	
Air Blank	0.000	01:00	
Subject Sample	# 1	01:00	Subject Test Refused
Breath Volume	0.000	01:00	
Air Blank	0.000	01:01	
Control Test	0.080	01:01	
Air Blank	0.000	01:02	
Diagnostics Check	OK	01:02	

Data Download Date/Time: 06/08/2018 13:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUFF,ZACHARY,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: HUFF,ZACHARY
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 05/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: IRIZARRY

DATE/TIME OF TEST: 05/11/2018 16:56:42

Diagnostics Check	OK	16:50
Air Blank	0.000	16:50
Control Test	0.079	16:50
Air Blank	0.000	16:51
Subject Sample # 1	0.000	16:52
Breath Volume	4.335	16:52
Air Blank	0.000	16:52
Air Blank	0.000	16:54
Subject Sample # 2	0.000	16:55
Breath Volume	3.734	16:55
Air Blank	0.000	16:55
Control Test	0.078	16:56
Air Blank	0.000	16:56
Diagnostics Check	OK	16:56

Data Download Date/Time: 06/08/2018 13:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SORICE, DOMINIC, M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: GOEDE, BRANDON
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 05/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 15:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:49

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 05/18/2018 16:16:08

Diagnostics Check	OK	16:10
Air Blank	0.000	16:10
Control Test	0.078	16:10
Air Blank	0.000	16:11
Subject Sample # 1	0.000	16:11
Breath Volume	3.957	16:11
Air Blank	0.000	16:12
Air Blank	0.000	16:14
Subject Sample # 2	0.000	16:14
Breath Volume	3.062	16:14
Air Blank	0.000	16:15
Control Test	0.078	16:15
Air Blank	0.000	16:16
Diagnostics Check	OK	16:16

Data Download Date/Time: 06/08/2018 13:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JASIUKIEWICZ,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 05/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: COLLADO

DATE/TIME OF TEST: 05/19/2018 02:55:00

Diagnostics Check	OK	02:51	
Air Blank	0.000	02:52	
Control Test	0.079	02:52	
Air Blank	0.000	02:52	
Subject Sample	# 1	02:53	Subject Test Refused
Breath Volume	0.000	02:53	
Air Blank	0.000	02:54	
Control Test	0.078	02:54	
Air Blank	0.000	02:54	
Diagnostics Check	OK	02:55	

Data Download Date/Time: 06/08/2018 13:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEAWARD,MORGANA,A
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SEAWARD,MORGANA
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 05/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:37

SUBJECT'S LAST NAME: YORK

DATE/TIME OF TEST: 05/20/2018 20:09:14

Diagnostics Check	OK	19:57	
Air Blank	0.000	19:57	
Control Test	0.079	19:58	
Air Blank	0.000	19:58	
Subject Sample	# 1 0.252	20:01	Volume Not Met
Breath Volume	0.085	20:01	
Air Blank	0.000	20:02	
Air Blank	0.000	20:04	
Subject Sample	# 2 0.262	20:07	Volume Not Met
Breath Volume	0.378	20:07	
Air Blank	0.000	20:08	
Control Test	0.073	20:08	Control Outside Tolerance
Air Blank	0.000	20:09	

Data Download Date/Time: 06/08/2018 13:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUFF,ZACHARY,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: HUFF,ZACHARY
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 05/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 15:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:24

SUBJECT'S LAST NAME: HAMILTON

DATE/TIME OF TEST: 05/25/2018 15:54:19

Diagnostics Check	OK	15:47
Air Blank	0.000	15:47
Control Test	0.079	15:47
Air Blank	0.000	15:48
Subject Sample # 1	0.000	15:49
Breath Volume	1.621	15:49
Air Blank	0.000	15:49
Air Blank	0.000	15:51
Subject Sample # 2	0.000	15:52
Breath Volume	1.535	15:52
Air Blank	0.000	15:53
Control Test	0.078	15:53
Air Blank	0.000	15:54
Diagnostics Check	OK	15:54

Data Download Date/Time: 06/08/2018 13:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEA,JOHN,D
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SHEA,JOHN
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 05/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: GROSCHEL

DATE/TIME OF TEST: 05/25/2018 23:32:05

Diagnostics Check	OK	23:25
Air Blank	0.000	23:26
Control Test	0.078	23:26
Air Blank	0.000	23:27
Subject Sample # 1	0.104	23:27
Breath Volume	3.296	23:27
Air Blank	0.000	23:28
Air Blank	0.000	23:30
Subject Sample # 2	0.098	23:30
Breath Volume	2.406	23:30
Air Blank	0.000	23:31
Control Test	0.077	23:31
Air Blank	0.000	23:32
Diagnostics Check	OK	23:32

Data Download Date/Time: 06/08/2018 13:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUFF,ZACHARY,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: HUFF,ZACHARY
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 05/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: REESE

DATE/TIME OF TEST: 05/26/2018 00:04:34

Diagnostics Check	OK	23:55
Air Blank	0.000	23:55
Control Test	0.078	23:55
Air Blank	0.000	23:56
Subject Sample # 2	0.136	23:59
Breath Volume	1.636	23:59
Air Blank	0.000	00:00
Air Blank	0.000	00:01
Subject Sample # 1	0.136	00:02
Breath Volume	2.449	00:02
Air Blank	0.000	00:03
Control Test	0.077	00:03
Air Blank	0.000	00:04
Diagnostics Check	OK	00:04

Data Download Date/Time: 06/08/2018 13:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOYLE,THOMAS,E
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: BOYLE,THOMAS
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 05/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: KRAFT

DATE/TIME OF TEST: 05/26/2018 01:33:54

Diagnostics Check	OK	01:27
Air Blank	0.000	01:28
Control Test	0.078	01:28
Air Blank	0.000	01:28
Subject Sample # 1	0.134	01:29
Breath Volume	2.906	01:29
Air Blank	0.000	01:30
Air Blank	0.000	01:31
Subject Sample # 2	0.130	01:32
Breath Volume	2.140	01:32
Air Blank	0.000	01:32
Control Test	0.077	01:33
Air Blank	0.000	01:33
Diagnostics Check	OK	01:33

Data Download Date/Time: 06/08/2018 13:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEA,JOHN,D
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SHEA,JOHN
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 05/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: ELLZEY

DATE/TIME OF TEST: 05/26/2018 01:42:13

Diagnostics Check	OK	01:39	
Air Blank	0.000	01:39	
Control Test	0.077	01:40	
Air Blank	0.000	01:40	
Subject Sample	# 1	01:40	Subject Test Refused
Breath Volume	0.000	01:40	
Air Blank	0.000	01:41	
Control Test	0.077	01:41	
Air Blank	0.000	01:42	
Diagnostics Check	OK	01:42	

Data Download Date/Time: 06/08/2018 13:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEA,JOHN,D
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SHEA,JOHN
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 05/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: JEFFERSON

DATE/TIME OF TEST: 05/26/2018 04:24:01

Diagnostics Check	OK	04:21	
Air Blank	0.000	04:21	
Control Test	0.078	04:21	
Air Blank	0.000	04:22	
Subject Sample	# 1	04:22	Subject Test Refused
Breath Volume	0.000	04:22	
Air Blank	0.000	04:23	
Control Test	0.077	04:23	
Air Blank	0.000	04:23	
Diagnostics Check	OK	04:24	

Data Download Date/Time: 06/08/2018 13:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOYLE,THOMAS,E
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: BOYLE,THOMAS
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 05/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 05/27/2018 02:04:08

Diagnostics Check	OK	01:57
Air Blank	0.000	01:58
Control Test	0.079	01:58
Air Blank	0.000	01:59
Subject Sample # 1	0.169	01:59
Breath Volume	1.714	01:59
Air Blank	0.000	02:00
Air Blank	0.000	02:02
Subject Sample # 2	0.176	02:02
Breath Volume	1.300	02:02
Air Blank	0.000	02:03
Control Test	0.077	02:03
Air Blank	0.000	02:04
Diagnostics Check	OK	02:04

Data Download Date/Time: 06/08/2018 13:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSEN,STEPHEN,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: CZOP,BRYAN
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 05/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: LONG

DATE/TIME OF TEST: 05/28/2018 02:44:24

Diagnostics Check	OK	02:41	
Air Blank	0.000	02:42	
Control Test	0.078	02:42	
Air Blank	0.000	02:42	
Subject Sample	# 1	02:43	Subject Test Refused
Breath Volume	0.000	02:43	
Air Blank	0.000	02:43	
Control Test	0.077	02:43	
Air Blank	0.000	02:44	
Diagnostics Check	OK	02:44	

Data Download Date/Time: 06/08/2018 13:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 05/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:26

SUBJECT'S LAST NAME: GAUDIELLO

DATE/TIME OF TEST: 05/30/2018 20:49:50

Diagnostics Check	OK	20:46	
Air Blank	0.000	20:47	
Control Test	0.078	20:47	
Air Blank	0.000	20:48	
Subject Sample	# 1	20:48	Subject Test Refused
Breath Volume	0.000	20:48	
Air Blank	0.000	20:48	
Control Test	0.076	20:49	
Air Blank	0.000	20:49	
Diagnostics Check	OK	20:49	

Data Download Date/Time: 06/08/2018 13:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEGERSMITH,DANIEL,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,DANIEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: DENNIS

DATE/TIME OF TEST: 05/03/2018 19:57:14

Diagnostics Check	OK	19:50
Air Blank	0.000	19:51
Control Test	0.079	19:51
Air Blank	0.000	19:51
Subject Sample # 1	0.000	19:52
Breath Volume	2.093	19:52
Air Blank	0.000	19:53
Air Blank	0.000	19:55
Subject Sample # 2	0.000	19:55
Breath Volume	1.378	19:55
Air Blank	0.000	19:56
Control Test	0.079	19:56
Air Blank	0.000	19:57
Diagnostics Check	OK	19:57

Data Download Date/Time: 06/27/2018 21:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUGENT,DAVID,A
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: HURT,J
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: MCGUIRE

DATE/TIME OF TEST: 05/03/2018 21:14:01

Diagnostics Check	OK	21:06
Air Blank	0.000	21:06
Control Test	0.079	21:07
Air Blank	0.000	21:07
Subject Sample # 1	0.202	21:08
Breath Volume	1.937	21:08
Air Blank	0.000	21:09
Air Blank	0.000	21:10
Subject Sample # 2	0.206	21:12
Breath Volume	1.890	21:12
Air Blank	0.000	21:13
Control Test	0.079	21:13
Air Blank	0.000	21:13
Diagnostics Check	OK	21:14

Data Download Date/Time: 06/27/2018 21:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VENEZIA,STEVEN,L
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: VENEZIA,STEVEN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: SANFORD

DATE/TIME OF TEST: 05/05/2018 05:22:53

Diagnostics Check	OK	05:16
Air Blank	0.000	05:16
Control Test	0.080	05:16
Air Blank	0.000	05:17
Subject Sample # 1	0.240	05:18
Breath Volume	2.132	05:18
Air Blank	0.000	05:18
Air Blank	0.000	05:20
Subject Sample # 2	0.249	05:21
Breath Volume	1.996	05:21
Air Blank	0.000	05:21
Control Test	0.080	05:22
Air Blank	0.000	05:22
Diagnostics Check	OK	05:22

Data Download Date/Time: 06/27/2018 21:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEGERSMITH,DANIEL,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,DANIEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:51

SUBJECT'S LAST NAME: MARONEY

DATE/TIME OF TEST: 05/05/2018 19:29:45

Diagnostics Check	OK	19:22
Air Blank	0.000	19:23
Control Test	0.081	19:23
Air Blank	0.000	19:23
Subject Sample # 1	0.000	19:25
Breath Volume	2.062	19:25
Air Blank	0.000	19:25
Air Blank	0.000	19:27
Subject Sample # 2	0.000	19:28
Breath Volume	2.410	19:28
Air Blank	0.000	19:28
Control Test	0.081	19:29
Air Blank	0.000	19:29
Diagnostics Check	OK	19:29

Data Download Date/Time: 06/27/2018 21:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THEODORE,J
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: CONYERS,CARL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: FOX

DATE/TIME OF TEST: 05/07/2018 02:00:44

Diagnostics Check	OK	01:54
Air Blank	0.000	01:55
Control Test	0.080	01:55
Air Blank	0.000	01:55
Subject Sample # 1	0.000	01:56
Breath Volume	1.433	01:56
Air Blank	0.000	01:56
Air Blank	0.000	01:58
Subject Sample # 2	0.000	01:59
Breath Volume	1.628	01:59
Air Blank	0.000	01:59
Control Test	0.080	02:00
Air Blank	0.000	02:00
Diagnostics Check	OK	02:00

Data Download Date/Time: 06/27/2018 21:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THEODORE,J
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: BERG,BRITTANI
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 05/07/2018 23:39:21

Diagnostics Check	OK	23:32
Air Blank	0.000	23:33
Control Test	0.080	23:33
Air Blank	0.000	23:34
Subject Sample # 1	0.164	23:34
Breath Volume	2.800	23:34
Air Blank	0.000	23:35
Air Blank	0.000	23:37
Subject Sample # 2	0.166	23:37
Breath Volume	2.164	23:37
Air Blank	0.000	23:38
Control Test	0.079	23:38
Air Blank	0.000	23:39
Diagnostics Check	OK	23:39

Data Download Date/Time: 06/27/2018 21:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,CHAD,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: REED,CHAD
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: NEWBERRY

DATE/TIME OF TEST: 05/20/2018 00:00:55

Diagnostics Check	OK	23:54
Air Blank	0.000	23:54
Control Test	0.079	23:54
Air Blank	0.000	23:55
Subject Sample # 1	0.000	23:56
Breath Volume	2.164	23:56
Air Blank	0.000	23:56
Air Blank	0.000	23:58
Subject Sample # 2	0.000	23:59
Breath Volume	2.058	23:59
Air Blank	0.000	23:59
Control Test	0.079	00:00
Air Blank	0.000	00:00
Diagnostics Check	OK	00:00

Data Download Date/Time: 06/27/2018 21:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,CHAD,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: HURT,JONATHAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: SPISSAK

DATE/TIME OF TEST: 05/24/2018 03:07:44

Diagnostics Check	OK	03:01
Air Blank	0.000	03:01
Control Test	0.079	03:01
Air Blank	0.000	03:02
Subject Sample # 1	0.039	03:03
Breath Volume	3.027	03:03
Air Blank	0.000	03:03
Air Blank	0.000	03:05
Subject Sample # 2	0.037	03:06
Breath Volume	3.035	03:06
Air Blank	0.000	03:06
Control Test	0.078	03:07
Air Blank	0.000	03:07
Diagnostics Check	OK	03:07

Data Download Date/Time: 06/27/2018 21:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,CHAD,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: HURT,JONATHAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: MARSHALL

DATE/TIME OF TEST: 05/25/2018 03:10:56

Diagnostics Check	OK	03:02
Air Blank	0.000	03:02
Control Test	0.079	03:03
Air Blank	0.000	03:03
Subject Sample # 1	0.173	03:06
Breath Volume	1.308	03:06
Air Blank	0.000	03:06
Air Blank	0.000	03:08
Subject Sample # 2	0.168	03:09
Breath Volume	1.320	03:09
Air Blank	0.000	03:09
Control Test	0.077	03:10
Air Blank	0.000	03:10
Diagnostics Check	OK	03:10

Data Download Date/Time: 06/27/2018 21:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,CHAD,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,DANIEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: HILLYER

DATE/TIME OF TEST: 05/25/2018 22:58:12

Diagnostics Check	OK	22:51
Air Blank	0.000	22:52
Control Test	0.078	22:52
Air Blank	0.000	22:53
Subject Sample # 1	0.020	22:53
Breath Volume	3.515	22:53
Air Blank	0.000	22:54
Air Blank	0.000	22:56
Subject Sample # 2	0.016	22:56
Breath Volume	3.578	22:56
Air Blank	0.000	22:57
Control Test	0.078	22:57
Air Blank	0.000	22:58
Diagnostics Check	OK	22:58

Data Download Date/Time: 06/27/2018 21:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,CHAD,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: HURT,JONATHAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: LIENHARD

DATE/TIME OF TEST: 05/28/2018 01:15:14

Diagnostics Check	OK	01:07
Air Blank	0.000	01:07
Control Test	0.079	01:08
Air Blank	0.000	01:08
Subject Sample # 1	0.061	01:10
Breath Volume	3.343	01:10
Air Blank	0.000	01:10
Air Blank	0.000	01:12
Subject Sample # 2	0.060	01:13
Breath Volume	3.242	01:13
Air Blank	0.000	01:14
Control Test	0.078	01:14
Air Blank	0.000	01:15
Diagnostics Check	OK	01:15

Data Download Date/Time: 06/27/2018 21:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,CHAD,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,DANIEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:11

SUBJECT'S LAST NAME: FRAZIER

DATE/TIME OF TEST: 05/29/2018 20:38:58

Diagnostics Check	OK	20:32
Air Blank	0.000	20:32
Control Test	0.079	20:32
Air Blank	0.000	20:33
Subject Sample # 1	0.080	20:34
Breath Volume	2.210	20:34
Air Blank	0.000	20:35
Air Blank	0.000	20:36
Subject Sample # 2	0.077	20:37
Breath Volume	1.953	20:37
Air Blank	0.000	20:38
Control Test	0.079	20:38
Air Blank	0.000	20:38
Diagnostics Check	OK	20:38

Data Download Date/Time: 06/27/2018 21:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: POCHE

DATE/TIME OF TEST: 05/01/2018 03:10:45

Diagnostics Check	OK	03:04
Air Blank	0.000	03:04
Control Test	0.082	03:04
Air Blank	0.000	03:05
Subject Sample # 1	0.231	03:06
Breath Volume	1.933	03:06
Air Blank	0.000	03:07
Air Blank	0.000	03:08
Subject Sample # 2	0.224	03:09
Breath Volume	1.515	03:09
Air Blank	0.000	03:09
Control Test	0.080	03:10
Air Blank	0.000	03:10
Diagnostics Check	OK	03:10

Data Download Date/Time: 05/01/2018 15:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNSON,G
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 05/02/2018 00:50:00

Diagnostics Check	OK	00:43
Air Blank	0.000	00:44
Control Test	0.081	00:44
Air Blank	0.000	00:45
Subject Sample # 1	0.102	00:45
Breath Volume	1.960	00:45
Air Blank	0.000	00:46
Air Blank	0.000	00:48
Subject Sample # 2	0.104	00:48
Breath Volume	2.179	00:48
Air Blank	0.000	00:49
Control Test	0.080	00:49
Air Blank	0.000	00:49
Diagnostics Check	OK	00:50

Data Download Date/Time: 05/30/2018 15:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: QUACH

DATE/TIME OF TEST: 05/02/2018 04:02:02

Diagnostics Check	OK	03:55
Air Blank	0.000	03:56
Control Test	0.082	03:56
Air Blank	0.000	03:56
Subject Sample # 1	0.219	03:57
Breath Volume	1.433	03:57
Air Blank	0.000	03:58
Air Blank	0.000	04:00
Subject Sample # 2	0.213	04:00
Breath Volume	2.046	04:00
Air Blank	0.000	04:01
Control Test	0.080	04:01
Air Blank	0.000	04:01
Diagnostics Check	OK	04:02

Data Download Date/Time: 05/30/2018 15:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DIAZ,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 05/02/2018 23:00:54

Diagnostics Check	OK	22:54
Air Blank	0.000	22:55
Control Test	0.082	22:55
Air Blank	0.000	22:56
Subject Sample # 1	0.210	22:56
Breath Volume	2.117	22:56
Air Blank	0.000	22:57
Air Blank	0.000	22:59
Subject Sample # 2	0.207	22:59
Breath Volume	1.855	22:59
Air Blank	0.000	22:59
Control Test	0.081	23:00
Air Blank	0.000	23:00
Diagnostics Check	OK	23:00

Data Download Date/Time: 05/30/2018 15:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: CARREON

DATE/TIME OF TEST: 05/03/2018 02:58:55

Diagnostics Check	OK	02:52
Air Blank	0.000	02:53
Control Test	0.081	02:53
Air Blank	0.000	02:54
Subject Sample # 1	0.109	02:54
Breath Volume	3.195	02:54
Air Blank	0.000	02:55
Air Blank	0.000	02:57
Subject Sample # 2	0.108	02:57
Breath Volume	2.808	02:57
Air Blank	0.000	02:57
Control Test	0.081	02:58
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

Data Download Date/Time: 05/30/2018 15:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY, SHANON, E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MARRERO, J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: POUSLAND

DATE/TIME OF TEST: 05/04/2018 00:55:48

Diagnostics Check	OK	00:49
Air Blank	0.000	00:50
Control Test	0.082	00:50
Air Blank	0.000	00:51
Subject Sample # 1	0.000	00:51
Breath Volume	2.996	00:51
Air Blank	0.000	00:51
Air Blank	0.000	00:53
Subject Sample # 2	0.000	00:54
Breath Volume	2.992	00:54
Air Blank	0.000	00:54
Control Test	0.081	00:55
Air Blank	0.000	00:55
Diagnostics Check	OK	00:55

Data Download Date/Time: 05/30/2018 15:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY, SHANON, E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE, J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: SATTERFIELD

DATE/TIME OF TEST: 05/04/2018 03:43:02

Diagnostics Check	OK	03:40	
Air Blank	0.000	03:40	
Control Test	0.081	03:40	
Air Blank	0.000	03:41	
Subject Sample	# 1	03:41	Subject Test Refused
Breath Volume	0.000	03:41	
Air Blank	0.000	03:42	
Control Test	0.081	03:42	
Air Blank	0.000	03:42	
Diagnostics Check	OK	03:43	

Data Download Date/Time: 05/30/2018 15:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY, SHANON, E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GOODE, D
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 12:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: LEANG

DATE/TIME OF TEST: 05/05/2018 03:08:23

Diagnostics Check	OK	03:02
Air Blank	0.000	03:02
Control Test	0.080	03:03
Air Blank	0.000	03:03
Subject Sample # 1	0.218	03:03
Breath Volume	1.714	03:03
Air Blank	0.000	03:04
Air Blank	0.000	03:06
Subject Sample # 2	0.224	03:06
Breath Volume	1.894	03:06
Air Blank	0.000	03:07
Control Test	0.080	03:07
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

Data Download Date/Time: 05/30/2018 15:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY, SHANON, E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES, J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: SARNELLE

DATE/TIME OF TEST: 05/05/2018 23:56:52

Diagnostics Check	OK	23:54	
Air Blank	0.000	23:54	
Control Test	0.081	23:54	
Air Blank	0.000	23:55	
Subject Sample	# 1	23:55	Subject Test Refused
Breath Volume	0.000	23:55	
Air Blank	0.000	23:55	
Control Test	0.081	23:56	
Air Blank	0.000	23:56	
Diagnostics Check	OK	23:56	

Data Download Date/Time: 05/30/2018 15:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY, SHANON, E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE, J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: CARDINALE

DATE/TIME OF TEST: 05/06/2018 04:33:16

Diagnostics Check	OK	04:27
Air Blank	0.000	04:27
Control Test	0.081	04:27
Air Blank	0.000	04:28
Subject Sample # 1	0.143	04:28
Breath Volume	2.019	04:28
Air Blank	0.000	04:29
Air Blank	0.000	04:31
Subject Sample # 2	0.135	04:31
Breath Volume	2.519	04:31
Air Blank	0.000	04:32
Control Test	0.081	04:32
Air Blank	0.000	04:33
Diagnostics Check	OK	04:33

Data Download Date/Time: 05/30/2018 15:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FILBERT, E
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: BROWN-MONTGOMERY

DATE/TIME OF TEST: 05/07/2018 01:46:34

Diagnostics Check	OK	01:39
Air Blank	0.000	01:40
Control Test	0.080	01:40
Air Blank	0.000	01:41
Subject Sample # 1	0.160	01:41
Breath Volume	2.117	01:41
Air Blank	0.000	01:42
Air Blank	0.000	01:44
Subject Sample # 2	0.165	01:45
Breath Volume	1.867	01:45
Air Blank	0.000	01:45
Control Test	0.080	01:45
Air Blank	0.000	01:46
Diagnostics Check	OK	01:46

Data Download Date/Time: 05/30/2018 15:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY, SHANON, E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MORGAN, J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: OROSZ

DATE/TIME OF TEST: 05/07/2018 23:49:32

Diagnostics Check	OK	23:43
Air Blank	0.000	23:44
Control Test	0.081	23:44
Air Blank	0.000	23:44
Subject Sample # 1	0.100	23:45
Breath Volume	2.289	23:45
Air Blank	0.000	23:45
Air Blank	0.000	23:47
Subject Sample # 2	0.101	23:47
Breath Volume	2.132	23:47
Air Blank	0.000	23:48
Control Test	0.081	23:48
Air Blank	0.000	23:49
Diagnostics Check	OK	23:49

Data Download Date/Time: 05/30/2018 15:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN, C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: TOOMEY

DATE/TIME OF TEST: 05/08/2018 00:48:56

Diagnostics Check	OK	00:41
Air Blank	0.000	00:41
Control Test	0.080	00:41
Air Blank	0.000	00:42
Subject Sample # 1	0.181	00:44
Breath Volume	1.593	00:44
Air Blank	0.000	00:44
Air Blank	0.000	00:46
Subject Sample # 2	0.180	00:47
Breath Volume	2.625	00:47
Air Blank	0.000	00:48
Control Test	0.080	00:48
Air Blank	0.000	00:48
Diagnostics Check	OK	00:48

Data Download Date/Time: 05/30/2018 15:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY, SHANON, E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE, J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: WESLEY

DATE/TIME OF TEST: 05/08/2018 03:43:39

Diagnostics Check	OK	03:40	
Air Blank	0.000	03:41	
Control Test	0.081	03:41	
Air Blank	0.000	03:42	
Subject Sample	# 1	03:42	Subject Test Refused
Breath Volume	0.000	03:42	
Air Blank	0.000	03:42	
Control Test	0.080	03:43	
Air Blank	0.000	03:43	
Diagnostics Check	OK	03:43	

Data Download Date/Time: 05/30/2018 15:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,D,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MEYERS,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 11:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:15

SUBJECT'S LAST NAME: SINNOTT

DATE/TIME OF TEST: 05/11/2018 13:44:10

Diagnostics Check	OK	13:37
Air Blank	0.000	13:37
Control Test	0.081	13:38
Air Blank	0.000	13:38
Subject Sample # 1	0.092	13:39
Breath Volume	2.140	13:39
Air Blank	0.000	13:39
Air Blank	0.000	13:41
Subject Sample # 2	0.089	13:42
Breath Volume	2.082	13:42
Air Blank	0.000	13:43
Control Test	0.080	13:43
Air Blank	0.000	13:44
Diagnostics Check	OK	13:44

Data Download Date/Time: 05/30/2018 15:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AUGUSTINE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: CALLE

DATE/TIME OF TEST: 05/13/2018 01:10:00

Diagnostics Check	OK	01:07	
Air Blank	0.000	01:07	
Control Test	0.080	01:07	
Air Blank	0.000	01:08	
Subject Sample	# 1	01:08	Subject Test Refused
Breath Volume	0.000	01:08	
Air Blank	0.000	01:09	
Control Test	0.081	01:09	
Air Blank	0.000	01:09	
Diagnostics Check	OK	01:10	

Data Download Date/Time: 05/30/2018 15:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AUGUSTINE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: MARRERO ECHEVARRIA

DATE/TIME OF TEST: 05/13/2018 03:40:35

Diagnostics Check	OK	03:34
Air Blank	0.000	03:34
Control Test	0.081	03:35
Air Blank	0.000	03:35
Subject Sample # 1	0.000	03:36
Breath Volume	1.925	03:36
Air Blank	0.000	03:36
Air Blank	0.000	03:38
Subject Sample # 2	0.000	03:39
Breath Volume	1.414	03:39
Air Blank	0.000	03:39
Control Test	0.081	03:39
Air Blank	0.000	03:40
Diagnostics Check	OK	03:40

Data Download Date/Time: 05/30/2018 15:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AUGUSTINE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HYSMITH,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:11

SUBJECT'S LAST NAME: PROM

DATE/TIME OF TEST: 05/13/2018 05:38:17

Diagnostics Check	OK	05:32
Air Blank	0.000	05:32
Control Test	0.081	05:32
Air Blank	0.000	05:33
Subject Sample # 1	0.120	05:33
Breath Volume	1.382	05:33
Air Blank	0.000	05:34
Air Blank	0.000	05:36
Subject Sample # 2	0.121	05:36
Breath Volume	1.320	05:36
Air Blank	0.000	05:37
Control Test	0.081	05:37
Air Blank	0.000	05:38
Diagnostics Check	OK	05:38

Data Download Date/Time: 05/30/2018 15:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AUGUSTINE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHOWDHURY,P
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:55

SUBJECT'S LAST NAME: HORAN

DATE/TIME OF TEST: 05/13/2018 06:18:36

Diagnostics Check	OK	06:15	
Air Blank	0.000	06:16	
Control Test	0.081	06:16	
Air Blank	0.000	06:16	
Subject Sample	# 1	06:17	Subject Test Refused
Breath Volume	0.000	06:17	
Air Blank	0.000	06:17	
Control Test	0.081	06:17	
Air Blank	0.000	06:18	
Diagnostics Check	OK	06:18	

Data Download Date/Time: 05/30/2018 15:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,D,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SCARBOROUGH,A
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 09:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:09

SUBJECT'S LAST NAME: TOVAR

DATE/TIME OF TEST: 05/14/2018 12:33:52

Diagnostics Check	OK	12:29	
Air Blank	0.000	12:30	
Control Test	0.081	12:30	
Air Blank	0.000	12:31	
Subject Sample	# 1	12:32	Subject Test Refused
Breath Volume	0.000	12:32	
Air Blank	0.000	12:32	
Control Test	0.080	12:33	
Air Blank	0.000	12:33	
Diagnostics Check	OK	12:33	

Data Download Date/Time: 05/30/2018 15:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,D,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SCARBROUGH,A
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 09:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:09

SUBJECT'S LAST NAME: TOVAR

DATE/TIME OF TEST: 05/14/2018 12:38:33

Diagnostics Check	OK	12:35	
Air Blank	0.000	12:35	
Control Test	0.080	12:36	
Air Blank	0.000	12:36	
Subject Sample	# 1	12:36	Slope Not Met
Breath Volume	1.062	12:36	
Air Blank	0.000	12:37	
Control Test	0.080	12:37	
Air Blank	0.000	12:38	
Diagnostics Check	OK	12:38	

Data Download Date/Time: 05/30/2018 15:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,D,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SCARBROUGH,A
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 09:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:36

SUBJECT'S LAST NAME: TOVAR

DATE/TIME OF TEST: 05/14/2018 13:03:14

Diagnostics Check	OK	12:56
Air Blank	0.000	12:56
Control Test	0.080	12:57
Air Blank	0.000	12:57
Subject Sample # 1	0.236	12:58
Breath Volume	1.464	12:58
Air Blank	0.000	12:59
Air Blank	0.000	13:00
Subject Sample # 2	0.223	13:01
Breath Volume	1.421	13:01
Air Blank	0.000	13:02
Control Test	0.080	13:02
Air Blank	0.000	13:03
Diagnostics Check	OK	13:03

Data Download Date/Time: 05/30/2018 15:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AUGUSTINE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 05/15/2018 01:00:56

Diagnostics Check	OK	00:57	
Air Blank	0.000	00:57	
Control Test	0.081	00:58	
Air Blank	0.000	00:58	
Subject Sample	# 1	00:59	Subject Test Refused
Breath Volume	0.000	00:59	
Air Blank	0.000	01:00	
Control Test	0.081	01:00	
Air Blank	0.000	01:00	
Diagnostics Check	OK	01:00	

Data Download Date/Time: 05/30/2018 15:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MEWKIRK,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:22

SUBJECT'S LAST NAME: CHIEVES

DATE/TIME OF TEST: 05/16/2018 05:49:03

Diagnostics Check	OK	05:43
Air Blank	0.000	05:43
Control Test	0.081	05:43
Air Blank	0.000	05:44
Subject Sample # 1	0.000	05:44
Breath Volume	1.218	05:44
Air Blank	0.000	05:45
Air Blank	0.000	05:47
Subject Sample # 2	0.000	05:47
Breath Volume	1.988	05:47
Air Blank	0.000	05:48
Control Test	0.081	05:48
Air Blank	0.000	05:48
Diagnostics Check	OK	05:49

Data Download Date/Time: 05/30/2018 15:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MORGAN,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:43

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 05/16/2018 07:07:16

Diagnostics Check	OK	07:04	
Air Blank	0.000	07:04	
Control Test	0.081	07:05	
Air Blank	0.000	07:05	
Subject Sample	# 1	07:05	Subject Test Refused
Breath Volume	0.000	07:05	
Air Blank	0.000	07:06	
Control Test	0.081	07:06	
Air Blank	0.000	07:07	
Diagnostics Check	OK	07:07	

Data Download Date/Time: 05/30/2018 15:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: WASHINGTON

DATE/TIME OF TEST: 05/17/2018 01:40:39

Diagnostics Check	OK	01:37	
Air Blank	0.000	01:38	
Control Test	0.081	01:38	
Air Blank	0.000	01:39	
Subject Sample	# 1	01:39	Subject Test Refused
Breath Volume	0.000	01:39	
Air Blank	0.000	01:39	
Control Test	0.082	01:40	
Air Blank	0.000	01:40	
Diagnostics Check	OK	01:40	

Data Download Date/Time: 05/30/2018 15:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES, J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: SHOPP

DATE/TIME OF TEST: 05/17/2018 04:32:40

Diagnostics Check	OK	04:19	
Air Blank	0.000	04:20	
Control Test	0.081	04:20	
Air Blank	0.000	04:21	
Subject Sample	# 1	04:24	No Sample Provided
Breath Volume	0.000	04:24	
Air Blank	0.000	04:24	
Air Blank	0.000	04:26	
Subject Sample	# 2	0.203	04:27
Breath Volume	1.785	04:27	
Air Blank	0.000	04:28	
Air Blank	0.000	04:30	
Subject Sample	# 3	0.189	04:31
Breath Volume	1.132	04:31	
Air Blank	0.000	04:31	
Control Test	0.080	04:32	
Air Blank	0.000	04:32	
Diagnostics Check	OK	04:32	

Data Download Date/Time: 05/30/2018 15:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GARCIA,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 05/18/2018 01:24:33

Diagnostics Check	OK	01:18
Air Blank	0.000	01:18
Control Test	0.081	01:19
Air Blank	0.000	01:19
Subject Sample # 1	0.261	01:20
Breath Volume	3.046	01:20
Air Blank	0.000	01:20
Air Blank	0.000	01:22
Subject Sample # 2	0.250	01:23
Breath Volume	2.964	01:23
Air Blank	0.000	01:23
Control Test	0.081	01:23
Air Blank	0.000	01:24
Diagnostics Check	OK	01:24

Data Download Date/Time: 05/30/2018 15:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 05/18/2018 02:26:05

Diagnostics Check	OK	02:23	
Air Blank	0.000	02:23	
Control Test	0.081	02:23	
Air Blank	0.000	02:24	
Subject Sample	# 1	02:24	Subject Test Refused
Breath Volume	0.000	02:24	
Air Blank	0.000	02:25	
Control Test	0.081	02:25	
Air Blank	0.000	02:26	
Diagnostics Check	OK	02:26	

Data Download Date/Time: 05/30/2018 15:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: MCGINNIS

DATE/TIME OF TEST: 05/18/2018 04:33:14

Diagnostics Check	OK	04:27
Air Blank	0.000	04:27
Control Test	0.080	04:27
Air Blank	0.000	04:28
Subject Sample # 1	0.100	04:28
Breath Volume	2.597	04:28
Air Blank	0.000	04:29
Air Blank	0.000	04:31
Subject Sample # 2	0.104	04:31
Breath Volume	3.085	04:31
Air Blank	0.000	04:32
Control Test	0.080	04:32
Air Blank	0.000	04:33
Diagnostics Check	OK	04:33

Data Download Date/Time: 05/30/2018 15:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: TARACHAND

DATE/TIME OF TEST: 05/19/2018 01:04:44

Diagnostics Check	OK	01:01	
Air Blank	0.000	01:02	
Control Test	0.080	01:02	
Air Blank	0.000	01:03	
Subject Sample	# 1	01:03	Subject Test Refused
Breath Volume	0.000	01:03	
Air Blank	0.000	01:03	
Control Test	0.081	01:04	
Air Blank	0.000	01:04	
Diagnostics Check	OK	01:04	

Data Download Date/Time: 05/30/2018 15:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: LO COCO

DATE/TIME OF TEST: 05/19/2018 03:05:16

Diagnostics Check	OK	02:59
Air Blank	0.000	02:59
Control Test	0.080	02:59
Air Blank	0.000	03:00
Subject Sample # 1	0.173	03:00
Breath Volume	3.347	03:00
Air Blank	0.000	03:01
Air Blank	0.000	03:03
Subject Sample # 2	0.183	03:03
Breath Volume	3.628	03:03
Air Blank	0.000	03:04
Control Test	0.080	03:04
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

Data Download Date/Time: 05/30/2018 15:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MAZZO,P
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: RILEY

DATE/TIME OF TEST: 05/19/2018 23:16:44

Diagnostics Check	OK	23:13	
Air Blank	0.000	23:14	
Control Test	0.081	23:14	
Air Blank	0.000	23:15	
Subject Sample	# 1	23:15	Subject Test Refused
Breath Volume	0.000	23:15	
Air Blank	0.000	23:15	
Control Test	0.080	23:16	
Air Blank	0.000	23:16	
Diagnostics Check	OK	23:16	

Data Download Date/Time: 05/30/2018 15:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 05/20/2018 04:25:58

Diagnostics Check	OK	04:23	
Air Blank	0.000	04:23	
Control Test	0.081	04:23	
Air Blank	0.000	04:24	
Subject Sample	# 1	04:24	Subject Test Refused
Breath Volume	0.000	04:24	
Air Blank	0.000	04:25	
Control Test	0.081	04:25	
Air Blank	0.000	04:25	
Diagnostics Check	OK	04:25	

Data Download Date/Time: 05/30/2018 15:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: DICKINSON

DATE/TIME OF TEST: 05/21/2018 00:49:45

Diagnostics Check	OK	00:43
Air Blank	0.000	00:44
Control Test	0.080	00:44
Air Blank	0.000	00:44
Subject Sample # 1	0.186	00:45
Breath Volume	2.699	00:45
Air Blank	0.000	00:45
Air Blank	0.000	00:47
Subject Sample # 2	0.185	00:48
Breath Volume	2.085	00:48
Air Blank	0.000	00:48
Control Test	0.081	00:49
Air Blank	0.000	00:49
Diagnostics Check	OK	00:49

Data Download Date/Time: 05/30/2018 15:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: COHILL,D
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: MONFORT

DATE/TIME OF TEST: 05/21/2018 04:21:05

Diagnostics Check	OK	04:15
Air Blank	0.000	04:15
Control Test	0.080	04:15
Air Blank	0.000	04:16
Subject Sample # 1	0.150	04:16
Breath Volume	3.019	04:16
Air Blank	0.000	04:17
Air Blank	0.000	04:19
Subject Sample # 2	0.154	04:19
Breath Volume	2.582	04:19
Air Blank	0.000	04:20
Control Test	0.081	04:20
Air Blank	0.000	04:21
Diagnostics Check	OK	04:21

Data Download Date/Time: 05/30/2018 15:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 05/23/2018 00:49:59

Diagnostics Check	OK	00:47	
Air Blank	0.000	00:47	
Control Test	0.081	00:47	
Air Blank	0.000	00:48	
Subject Sample	# 1	00:48	Subject Test Refused
Breath Volume	0.000	00:48	
Air Blank	0.000	00:49	
Control Test	0.081	00:49	
Air Blank	0.000	00:49	
Diagnostics Check	OK	00:49	

Data Download Date/Time: 05/30/2018 15:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CAMERON,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:25

SUBJECT'S LAST NAME: BOYINGTON

DATE/TIME OF TEST: 05/23/2018 05:49:02

Diagnostics Check	OK	05:46	
Air Blank	0.000	05:46	
Control Test	0.081	05:46	
Air Blank	0.000	05:47	
Subject Sample	# 1	05:47	Subject Test Refused
Breath Volume	0.000	05:47	
Air Blank	0.000	05:48	
Control Test	0.081	05:48	
Air Blank	0.000	05:48	
Diagnostics Check	OK	05:49	

Data Download Date/Time: 05/30/2018 15:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: VAZQUEZ-RAMIREZ

DATE/TIME OF TEST: 05/24/2018 03:17:38

Diagnostics Check	OK	03:10
Air Blank	0.000	03:10
Control Test	0.081	03:11
Air Blank	0.000	03:11
Subject Sample # 1	0.199	03:12
Breath Volume	1.957	03:12
Air Blank	0.000	03:13
Air Blank	0.000	03:15
Subject Sample # 2	0.197	03:16
Breath Volume	1.546	03:16
Air Blank	0.000	03:16
Control Test	0.080	03:17
Air Blank	0.000	03:17
Diagnostics Check	OK	03:17

Data Download Date/Time: 05/30/2018 15:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JENSEN,D
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:06

SUBJECT'S LAST NAME: HUNNICUTT

DATE/TIME OF TEST: 05/24/2018 05:32:55

Diagnostics Check	OK	05:26
Air Blank	0.000	05:27
Control Test	0.081	05:27
Air Blank	0.000	05:28
Subject Sample # 1	0.204	05:28
Breath Volume	4.125	05:28
Air Blank	0.000	05:29
Air Blank	0.000	05:31
Subject Sample # 2	0.202	05:31
Breath Volume	4.320	05:31
Air Blank	0.000	05:31
Control Test	0.080	05:32
Air Blank	0.000	05:32
Diagnostics Check	OK	05:32

Data Download Date/Time: 05/30/2018 15:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRYNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 05/25/2018 00:15:12

Diagnostics Check	OK	00:12	
Air Blank	0.000	00:12	
Control Test	0.080	00:13	
Air Blank	0.000	00:13	
Subject Sample	# 1	00:13	Subject Test Refused
Breath Volume	0.000	00:13	
Air Blank	0.000	00:14	
Control Test	0.081	00:14	
Air Blank	0.000	00:15	
Diagnostics Check	OK	00:15	

Data Download Date/Time: 05/30/2018 15:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE, J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: WOLFE

DATE/TIME OF TEST: 05/25/2018 03:51:12

Diagnostics Check	OK	03:44
Air Blank	0.000	03:45
Control Test	0.080	03:45
Air Blank	0.000	03:46
Subject Sample # 1	0.218	03:46
Breath Volume	4.378	03:46
Air Blank	0.000	03:47
Air Blank	0.000	03:49
Subject Sample # 2	0.222	03:49
Breath Volume	4.539	03:49
Air Blank	0.000	03:50
Control Test	0.081	03:50
Air Blank	0.000	03:51
Diagnostics Check	OK	03:51

Data Download Date/Time: 05/30/2018 15:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRYANT, R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 05/26/2018 03:37:09

Diagnostics Check	OK	03:28
Air Blank	0.000	03:28
Control Test	0.081	03:29
Air Blank	0.000	03:29
Subject Sample # 1	0.185	03:31
Breath Volume	1.429	03:31
Air Blank	0.000	03:32
Air Blank	0.000	03:34
Subject Sample # 2	0.181	03:35
Breath Volume	2.000	03:35
Air Blank	0.000	03:36
Control Test	0.081	03:36
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 05/30/2018 15:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY, SHANON, E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JENSEN, D
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:46

SUBJECT'S LAST NAME: GOLDEN

DATE/TIME OF TEST: 05/26/2018 05:12:55

Diagnostics Check	OK	05:06
Air Blank	0.000	05:07
Control Test	0.081	05:07
Air Blank	0.000	05:07
Subject Sample # 1	0.249	05:08
Breath Volume	1.726	05:08
Air Blank	0.000	05:08
Air Blank	0.000	05:10
Subject Sample # 2	0.234	05:11
Breath Volume	2.378	05:11
Air Blank	0.000	05:11
Control Test	0.081	05:12
Air Blank	0.000	05:12
Diagnostics Check	OK	05:12

Data Download Date/Time: 05/30/2018 15:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY, SHANON, E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HORTON, BREANNA
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:01

SUBJECT'S LAST NAME: HORTON

DATE/TIME OF TEST: 05/26/2018 23:25:35

Diagnostics Check	OK	23:22	
Air Blank	0.000	23:23	
Control Test	0.081	23:23	
Air Blank	0.000	23:23	
Subject Sample	# 1	23:24	Subject Test Refused
Breath Volume	0.000	23:24	
Air Blank	0.000	23:24	
Control Test	0.081	23:24	
Air Blank	0.000	23:25	
Diagnostics Check	OK	23:25	

Data Download Date/Time: 05/30/2018 15:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: DANIELS

DATE/TIME OF TEST: 05/27/2018 02:23:29

Diagnostics Check	OK	02:20	
Air Blank	0.000	02:21	
Control Test	0.081	02:21	
Air Blank	0.000	02:21	
Subject Sample	# 1	02:22	Subject Test Refused
Breath Volume	0.000	02:22	
Air Blank	0.000	02:22	
Control Test	0.080	02:22	
Air Blank	0.000	02:23	
Diagnostics Check	OK	02:23	

Data Download Date/Time: 05/30/2018 15:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY, SHANON, E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BUTLER, J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:47

SUBJECT'S LAST NAME: ARREGUIN

DATE/TIME OF TEST: 05/27/2018 06:15:04

Diagnostics Check	OK	06:08
Air Blank	0.000	06:09
Control Test	0.080	06:09
Air Blank	0.000	06:09
Subject Sample # 1	0.170	06:10
Breath Volume	2.359	06:10
Air Blank	0.000	06:10
Air Blank	0.000	06:12
Subject Sample # 2	0.180	06:13
Breath Volume	2.484	06:13
Air Blank	0.000	06:14
Control Test	0.080	06:14
Air Blank	0.000	06:14
Diagnostics Check	OK	06:15

Data Download Date/Time: 05/30/2018 15:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE, J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: PRIM

DATE/TIME OF TEST: 05/28/2018 00:01:52

Diagnostics Check	OK	23:53
Air Blank	0.000	23:53
Control Test	0.080	23:54
Air Blank	0.000	23:54
Subject Sample # 2	0.025	23:56
Breath Volume	1.128	23:56
Air Blank	0.000	23:57
Air Blank	0.000	23:59
Subject Sample # 1	0.021	00:00
Breath Volume	1.101	00:00
Air Blank	0.000	00:00
Control Test	0.081	00:01
Air Blank	0.000	00:01
Diagnostics Check	OK	00:01

Data Download Date/Time: 05/30/2018 15:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: Fiset, J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: STRATTON

DATE/TIME OF TEST: 05/28/2018 02:07:18

Diagnostics Check	OK	02:04	
Air Blank	0.000	02:04	
Control Test	0.081	02:05	
Air Blank	0.000	02:05	
Subject Sample	# 1	02:05	Subject Test Refused
Breath Volume	0.000	02:05	
Air Blank	0.000	02:06	
Control Test	0.080	02:06	
Air Blank	0.000	02:07	
Diagnostics Check	OK	02:07	

Data Download Date/Time: 05/30/2018 15:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY, SHANON, E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER, R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: AGUILAR

DATE/TIME OF TEST: 05/28/2018 03:31:57

Diagnostics Check	OK	03:29	
Air Blank	0.000	03:29	
Control Test	0.081	03:29	
Air Blank	0.000	03:30	
Subject Sample	# 1	03:30	Subject Test Refused
Breath Volume	0.000	03:30	
Air Blank	0.000	03:31	
Control Test	0.081	03:31	
Air Blank	0.000	03:31	
Diagnostics Check	OK	03:31	

Data Download Date/Time: 05/30/2018 15:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY, SHANON, E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GARCIA, J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: ROE

DATE/TIME OF TEST: 05/28/2018 04:23:49

Diagnostics Check	OK	04:20	
Air Blank	0.000	04:21	
Control Test	0.081	04:21	
Air Blank	0.000	04:22	
Subject Sample	# 1	04:22	Subject Test Refused
Breath Volume	0.000	04:22	
Air Blank	0.000	04:22	
Control Test	0.082	04:23	
Air Blank	0.000	04:23	
Diagnostics Check	OK	04:23	

Data Download Date/Time: 05/30/2018 15:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY, SHANON, E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GARCIA, J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: ADMIN
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: ROE

DATE/TIME OF TEST: 05/28/2018 04:54:47

Diagnostics Check	OK	04:48
Air Blank	0.000	04:48
Control Test	0.081	04:49
Air Blank	0.000	04:49
Subject Sample # 1	0.200	04:50
Breath Volume	2.246	04:50
Air Blank	0.000	04:50
Air Blank	0.000	04:52
Subject Sample # 2	0.215	04:53
Breath Volume	2.414	04:53
Air Blank	0.000	04:53
Control Test	0.081	04:54
Air Blank	0.000	04:54
Diagnostics Check	OK	04:54

Data Download Date/Time: 05/30/2018 15:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HAWKINS,A
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: GONSALVES

DATE/TIME OF TEST: 05/29/2018 03:20:25

Diagnostics Check	OK	03:17	
Air Blank	0.000	03:18	
Control Test	0.081	03:18	
Air Blank	0.000	03:18	
Subject Sample	# 1	03:19	Subject Test Refused
Breath Volume	0.000	03:19	
Air Blank	0.000	03:19	
Control Test	0.082	03:19	
Air Blank	0.000	03:20	
Diagnostics Check	OK	03:20	

Data Download Date/Time: 05/30/2018 15:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GARCIA,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: WILKINSON

DATE/TIME OF TEST: 05/29/2018 23:52:54

Diagnostics Check	OK	23:50	
Air Blank	0.000	23:50	
Control Test	0.081	23:50	
Air Blank	0.000	23:51	
Subject Sample	# 1	23:51	Subject Test Refused
Breath Volume	0.000	23:51	
Air Blank	0.000	23:51	
Control Test	0.081	23:52	
Air Blank	0.000	23:52	
Diagnostics Check	OK	23:52	

Data Download Date/Time: 05/30/2018 15:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: SHEHEE

DATE/TIME OF TEST: 05/30/2018 03:45:26

Diagnostics Check	OK	03:39
Air Blank	0.000	03:39
Control Test	0.081	03:40
Air Blank	0.000	03:40
Subject Sample # 1	0.146	03:41
Breath Volume	2.492	03:41
Air Blank	0.000	03:41
Air Blank	0.000	03:43
Subject Sample # 2	0.147	03:43
Breath Volume	2.496	03:43
Air Blank	0.000	03:44
Control Test	0.081	03:44
Air Blank	0.000	03:45
Diagnostics Check	OK	03:45

Data Download Date/Time: 05/30/2018 15:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: JORDAN

DATE/TIME OF TEST: 05/31/2018 01:07:52

Diagnostics Check	OK	01:01
Air Blank	0.000	01:02
Control Test	0.082	01:02
Air Blank	0.000	01:03
Subject Sample # 1	0.000	01:03
Breath Volume	5.257	01:03
Air Blank	0.000	01:04
Air Blank	0.000	01:06
Subject Sample # 2	0.000	01:06
Breath Volume	5.488	01:06
Air Blank	0.000	01:06
Control Test	0.082	01:07
Air Blank	0.000	01:07
Diagnostics Check	OK	01:07

Data Download Date/Time: 06/19/2018 15:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ARGUDO,L
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:17

SUBJECT'S LAST NAME: MYSTKOWSKI

DATE/TIME OF TEST: 05/31/2018 05:41:14

Diagnostics Check	OK	05:38	
Air Blank	0.000	05:38	
Control Test	0.080	05:39	
Air Blank	0.000	05:39	
Subject Sample	# 1	05:39	Subject Test Refused
Breath Volume	0.000	05:39	
Air Blank	0.000	05:40	
Control Test	0.081	05:40	
Air Blank	0.000	05:41	
Diagnostics Check	OK	05:41	

Data Download Date/Time: 06/19/2018 15:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

INSTRUMENT REGISTERED TO: HIALEAH PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CONSALVO, ESTEBAN, A

BREATH TEST OPERATOR AGENCY: Hialeah PD

ARRESTING OFFICER: AGUERO, RICHARD

ARRESTING OFFICER AGENCY: Hialeah PD

LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 712031

DRY GAS EXPIRY DATE: 02/21/2020

VIOLATION CODE: DUI

ARREST TIME: 02:50

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: AMEZQUITA

DATE/TIME OF TEST: 05/12/2018 05:17:32

Diagnostics Check	OK	05:14	
Air Blank	0.000	05:15	
Control Test	0.079	05:15	
Air Blank	0.000	05:15	
Subject Sample	# 1	05:16	Subject Test Refused
Breath Volume	0.000	05:16	
Air Blank	0.000	05:16	
Control Test	0.080	05:16	
Air Blank	0.000	05:17	
Diagnostics Check	OK	05:17	

Data Download Date/Time: 05/21/2018 10:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LARREA,S,
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: LARREA,S
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 07:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:17

SUBJECT'S LAST NAME: LEON

DATE/TIME OF TEST: 05/12/2018 08:40:47

Diagnostics Check	OK	08:37	
Air Blank	0.000	08:38	
Control Test	0.079	08:38	
Air Blank	0.000	08:39	
Subject Sample	# 1	08:39	Subject Test Refused
Breath Volume	0.000	08:39	
Air Blank	0.000	08:39	
Control Test	0.080	08:40	
Air Blank	0.000	08:40	
Diagnostics Check	OK	08:40	

Data Download Date/Time: 05/21/2018 10:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,EVI,
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: HERNADEZ,OSDANY
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 05/13/2018 01:35:01

Diagnostics Check	OK	01:31	
Air Blank	0.000	01:31	
Control Test	0.079	01:32	
Air Blank	0.000	01:32	
Subject Sample	# 1	01:33	Subject Test Refused
Breath Volume	0.000	01:33	
Air Blank	0.000	01:34	
Control Test	0.079	01:34	
Air Blank	0.000	01:34	
Diagnostics Check	OK	01:35	

Data Download Date/Time: 05/21/2018 10:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LARREA,S,
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: AGUERO,R
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 08:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:55

SUBJECT'S LAST NAME: HERNANDEZ MIRANDA

DATE/TIME OF TEST: 05/14/2018 09:25:22

Diagnostics Check	OK	09:16
Air Blank	0.000	09:17
Control Test	0.079	09:17
Air Blank	0.000	09:18
Subject Sample # 1	0.042	09:20
Breath Volume	1.722	09:20
Air Blank	0.000	09:21
Air Blank	0.000	09:23
Subject Sample # 2	0.036	09:23
Breath Volume	1.882	09:23
Air Blank	0.000	09:24
Control Test	0.077	09:24
Air Blank	0.000	09:25
Diagnostics Check	OK	09:25

Data Download Date/Time: 05/21/2018 10:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

INSTRUMENT REGISTERED TO: HIALEAH PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ARROJAS,GABRIEL,L

BREATH TEST OPERATOR AGENCY: Hialeah PD

ARRESTING OFFICER: ARIAS,ERNESTO

ARRESTING OFFICER AGENCY: Hialeah PD

LAST AGENCY INSPECTION: 05/21/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 712031

DRY GAS EXPIRY DATE: 02/21/2020

VIOLATION CODE: DUI

ARREST TIME: 02:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: ACOSTA

DATE/TIME OF TEST: 05/25/2018 03:16:30

Diagnostics Check	OK	03:10
Air Blank	0.000	03:10
Control Test	0.078	03:10
Air Blank	0.000	03:11
Subject Sample # 1	0.226	03:11
Breath Volume	1.433	03:11
Air Blank	0.000	03:12
Air Blank	0.000	03:14
Subject Sample # 2	0.227	03:14
Breath Volume	2.664	03:14
Air Blank	0.000	03:15
Control Test	0.078	03:15
Air Blank	0.000	03:16
Diagnostics Check	OK	03:16

Data Download Date/Time: 06/11/2018 10:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANO,A,
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: KELLY,D
ARRESTING OFFICER AGENCY: Opa Locka PD
LAST AGENCY INSPECTION: 05/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 18:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:54

SUBJECT'S LAST NAME: MARIN

DATE/TIME OF TEST: 05/25/2018 21:29:55

Diagnostics Check	OK	21:27	
Air Blank	0.000	21:27	
Control Test	0.079	21:27	
Air Blank	0.000	21:28	
Subject Sample	# 1	21:28	Subject Test Refused
Breath Volume	0.000	21:28	
Air Blank	0.000	21:28	
Control Test	0.079	21:29	
Air Blank	0.000	21:29	
Diagnostics Check	OK	21:29	

Data Download Date/Time: 06/11/2018 10:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

INSTRUMENT REGISTERED TO: HIALEAH PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CONSALVO,ESTEBAN,A

BREATH TEST OPERATOR AGENCY: Hialeah PD

ARRESTING OFFICER: VELOZ,ERIC

ARRESTING OFFICER AGENCY: Hialeah PD

LAST AGENCY INSPECTION: 05/21/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 712031

DRY GAS EXPIRY DATE: 02/21/2020

VIOLATION CODE: DUI

ARREST TIME: 02:50

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: HERNANDEZ GOMEZ

DATE/TIME OF TEST: 05/27/2018 04:13:38

Diagnostics Check	OK	04:10	
Air Blank	0.000	04:11	
Control Test	0.079	04:11	
Air Blank	0.000	04:12	
Subject Sample	# 1	04:12	Subject Test Refused
Breath Volume	0.000	04:12	
Air Blank	0.000	04:12	
Control Test	0.079	04:13	
Air Blank	0.000	04:13	
Diagnostics Check	OK	04:13	

Data Download Date/Time: 06/11/2018 10:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,AILEEN,B
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: CURBELO,DANIEL
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 05/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: HURTADO

DATE/TIME OF TEST: 05/29/2018 00:52:35

Diagnostics Check	OK	00:40	
Air Blank	0.000	00:40	
Control Test	0.079	00:41	
Air Blank	0.000	00:41	
Subject Sample	# 1	00:44	No Sample Provided
Breath Volume	0.000	00:44	
Air Blank	0.000	00:45	
Air Blank	0.000	00:47	
Subject Sample	# 2	0.094	00:47
Breath Volume	3.542	00:47	
Air Blank	0.000	00:48	
Air Blank	0.000	00:50	
Subject Sample	# 3	0.098	00:50
Breath Volume	2.796	00:50	
Air Blank	0.000	00:51	
Control Test	0.079	00:51	
Air Blank	0.000	00:52	
Diagnostics Check	OK	00:52	

Data Download Date/Time: 06/11/2018 10:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

INSTRUMENT REGISTERED TO: HIALEAH PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ARROJAS,GABRIEL,L

BREATH TEST OPERATOR AGENCY: Hialeah PD

ARRESTING OFFICER: IRIARTE,JOSE

ARRESTING OFFICER AGENCY: Hialeah PD

LAST AGENCY INSPECTION: 05/21/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 712031

DRY GAS EXPIRY DATE: 02/21/2020

VIOLATION CODE: DUI

ARREST TIME: 01:18

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: MURGADO

DATE/TIME OF TEST: 05/29/2018 03:17:11

Diagnostics Check	OK	03:14	
Air Blank	0.000	03:14	
Control Test	0.079	03:14	
Air Blank	0.000	03:15	
Subject Sample	# 1	03:15	Subject Test Refused
Breath Volume	0.000	03:15	
Air Blank	0.000	03:16	
Control Test	0.079	03:16	
Air Blank	0.000	03:17	
Diagnostics Check	OK	03:17	

Data Download Date/Time: 06/11/2018 10:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,G,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,G
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 04/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: ESTES

DATE/TIME OF TEST: 05/06/2018 00:07:52

Diagnostics Check	OK	00:01
Air Blank	0.000	00:01
Control Test	0.079	00:02
Air Blank	0.000	00:02
Subject Sample # 1	0.155	00:03
Breath Volume	1.703	00:03
Air Blank	0.000	00:03
Air Blank	0.000	00:05
Subject Sample # 2	0.157	00:06
Breath Volume	1.550	00:06
Air Blank	0.000	00:06
Control Test	0.079	00:07
Air Blank	0.000	00:07
Diagnostics Check	OK	00:07

Data Download Date/Time: 05/30/2018 11:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,G,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,G
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 04/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: ROEH

DATE/TIME OF TEST: 05/08/2018 22:41:33

Diagnostics Check	OK	22:38	
Air Blank	0.000	22:38	
Control Test	0.080	22:38	
Air Blank	0.000	22:39	
Subject Sample	# 1	22:39	Interferent Detect
Breath Volume	2.625	22:39	
Air Blank	0.000	22:40	
Control Test	0.080	22:40	
Air Blank	0.000	22:41	
Diagnostics Check	OK	22:41	

Data Download Date/Time: 05/30/2018 11:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,G,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,G
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 04/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: WALSH

DATE/TIME OF TEST: 05/18/2018 00:07:15

Diagnostics Check	OK	23:59
Air Blank	0.000	00:00
Control Test	0.078	00:00
Air Blank	0.000	00:01
Subject Sample # 1	0.143	00:02
Breath Volume	3.070	00:02
Air Blank	0.000	00:03
Air Blank	0.000	00:05
Subject Sample # 2	0.135	00:05
Breath Volume	1.351	00:05
Air Blank	0.000	00:06
Control Test	0.077	00:06
Air Blank	0.000	00:07
Diagnostics Check	OK	00:07

Data Download Date/Time: 05/30/2018 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,JEFFREY,A
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: THOMPSON,JEFFREY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 04/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: POOLE

DATE/TIME OF TEST: 05/21/2018 22:31:58

Diagnostics Check	OK	22:25
Air Blank	0.000	22:25
Control Test	0.079	22:26
Air Blank	0.000	22:26
Subject Sample # 1	0.000	22:27
Breath Volume	2.480	22:27
Air Blank	0.000	22:28
Air Blank	0.000	22:30
Subject Sample # 2	0.000	22:30
Breath Volume	1.675	22:30
Air Blank	0.000	22:31
Control Test	0.078	22:31
Air Blank	0.000	22:31
Diagnostics Check	OK	22:31

Data Download Date/Time: 05/30/2018 11:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,JEFFREY,A
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: THOMPSON,JEFFREY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 04/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:54

SUBJECT'S LAST NAME: FERRON

DATE/TIME OF TEST: 05/22/2018 22:21:26

Diagnostics Check	OK	22:16	
Air Blank	0.000	22:17	
Control Test	0.078	22:17	
Air Blank	0.000	22:18	
Subject Sample	# 1	22:20	Subject Test Refused
Breath Volume	0.000	22:20	
Air Blank	0.000	22:20	
Control Test	0.078	22:20	
Air Blank	0.000	22:21	
Diagnostics Check	OK	22:21	

Data Download Date/Time: 05/30/2018 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003196

INSTRUMENT REGISTERED TO: FWC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MCKEE,DONNIE,R

BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)

ARRESTING OFFICER: WARD,WILLIAM

ARRESTING OFFICER AGENCY: FFWCC (Ocala)

LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG629802

DRY GAS EXPIRY DATE: 10/24/2018

VIOLATION CODE: BUI

ARREST TIME: 18:47

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 05/05/2018 19:41:03

Diagnostics Check	OK	19:40	
Air Blank		19:40	RFI Detect
Air Blank	0.000	19:41	

Data Download Date/Time: 06/21/2018 11:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003196

INSTRUMENT REGISTERED TO: FWC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MCKEE,DONNIE,R

BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)

ARRESTING OFFICER: WARD,WILLIAM

ARRESTING OFFICER AGENCY: FFWCC (Ocala)

LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG629802

DRY GAS EXPIRY DATE: 10/24/2018

VIOLATION CODE: BUI

ARREST TIME: 18:27

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 05/05/2018 19:46:14

Diagnostics Check	OK	19:44	
Air Blank	0.000	19:45	
Control Test	0.072	19:45	Control Outside Tolerance
Air Blank	0.000	19:46	

Data Download Date/Time: 06/21/2018 11:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003196

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKEE,DONNIE,R
BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)
ARRESTING OFFICER: WARD,WILIAM
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: BUI
ARREST TIME: 18:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 05/05/2018 19:57:14

Diagnostics Check	OK	19:50
Air Blank	0.000	19:51
Control Test	0.076	19:51
Air Blank	0.000	19:52
Subject Sample # 1	0.190	19:52
Breath Volume	2.316	19:52
Air Blank	0.000	19:53
Air Blank	0.000	19:55
Subject Sample # 2	0.187	19:55
Breath Volume	3.316	19:55
Air Blank	0.000	19:56
Control Test	0.075	19:56
Air Blank	0.000	19:57
Diagnostics Check	OK	19:57

Data Download Date/Time: 06/21/2018 11:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003196

INSTRUMENT REGISTERED TO: FWC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MCKEE,DONNIE,R

BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)

ARRESTING OFFICER: MCKEE,DONNIE

ARRESTING OFFICER AGENCY: FFWCC (Ocala)

LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: Yes

DRY GAS STANDARD LOT #: AG629802

DRY GAS EXPIRY DATE: 10/24/2018

VIOLATION CODE: BUI

ARREST TIME: 16:58

GENDER: Female

OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: SPONDER

DATE/TIME OF TEST: 05/12/2018 18:23:38

Diagnostics Check	OK	18:14
Air Blank	0.000	18:15
Control Test	0.077	18:15
Air Blank	0.000	18:16
Subject Sample # 1	0.114	18:19
Breath Volume	1.976	18:19
Air Blank	0.000	18:19
Air Blank	0.000	18:21
Subject Sample # 2	0.116	18:22
Breath Volume	1.679	18:22
Air Blank	0.000	18:22
Control Test	0.078	18:23
Air Blank	0.000	18:23
Diagnostics Check	OK	18:23

Data Download Date/Time: 06/21/2018 11:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003197

INSTRUMENT REGISTERED TO: FWC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 477627
BREATH TEST OPERATOR: LAND,MATTHEW,A	DRY GAS EXPIRY DATE: 06/29/2019
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)	VIOLATION CODE: BUI
ARRESTING OFFICER: LAND,MATTHEW	ARREST TIME: 18:24
ARRESTING OFFICER AGENCY: FFWCC (Panama City)	GENDER: Male
LAST AGENCY INSPECTION: 05/09/2018	OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: PETTIGREW DATE/TIME OF TEST: 05/12/2018 19:54:00

Diagnostics Check	OK	19:47
Air Blank	0.000	19:48
Control Test	0.078	19:48
Air Blank	0.000	19:49
Subject Sample # 1	0.187	19:49
Breath Volume	2.187	19:49
Air Blank	0.000	19:50
Air Blank	0.000	19:51
Subject Sample # 2	0.182	19:52
Breath Volume	2.941	19:52
Air Blank	0.000	19:53
Control Test	0.078	19:53
Air Blank	0.000	19:53
Diagnostics Check	OK	19:54

Data Download Date/Time: 06/13/2018 18:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003197

INSTRUMENT REGISTERED TO: FWC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 477627
BREATH TEST OPERATOR: HAHR,DANIEL,E	DRY GAS EXPIRY DATE: 06/29/2019
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)	VIOLATION CODE: BUI
ARRESTING OFFICER: HAHR,DANIEL	ARREST TIME: 00:12
ARRESTING OFFICER AGENCY: FFWCC (Panama City)	GENDER: Male
LAST AGENCY INSPECTION: 05/09/2018	OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: WHITE DATE/TIME OF TEST: 05/27/2018 01:17:59

Diagnostics Check	OK	01:14	
Air Blank	0.000	01:15	
Control Test	0.081	01:15	
Air Blank	0.000	01:15	
Subject Sample	# 1	01:16	Subject Test Refused
Breath Volume	0.000	01:16	
Air Blank	0.000	01:17	
Control Test	0.078	01:17	
Air Blank	0.000	01:17	
Diagnostics Check	OK	01:17	

Data Download Date/Time: 06/13/2018 18:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,BARRY,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: NIXON,BARRY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 04/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:07

SUBJECT'S LAST NAME: CRANE

DATE/TIME OF TEST: 05/05/2018 05:34:33

Diagnostics Check	OK	05:28
Air Blank	0.000	05:28
Control Test	0.077	05:28
Air Blank	0.000	05:29
Subject Sample # 1	0.100	05:30
Breath Volume	1.847	05:30
Air Blank	0.000	05:30
Air Blank	0.000	05:32
Subject Sample # 2	0.092	05:33
Breath Volume	1.425	05:33
Air Blank	0.000	05:33
Control Test	0.077	05:33
Air Blank	0.000	05:34
Diagnostics Check	OK	05:34

Data Download Date/Time: 05/29/2018 17:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVARADO,ANGEL,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ALVARADO,ANGEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: JEFFERSON

DATE/TIME OF TEST: 05/05/2018 21:53:52

Diagnostics Check	OK	21:46
Air Blank	0.000	21:46
Control Test	0.078	21:47
Air Blank	0.000	21:47
Subject Sample # 1	0.115	21:48
Breath Volume	1.855	21:48
Air Blank	0.000	21:49
Air Blank	0.000	21:51
Subject Sample # 2	0.124	21:52
Breath Volume	2.359	21:52
Air Blank	0.000	21:52
Control Test	0.076	21:53
Air Blank	0.000	21:53
Diagnostics Check	OK	21:53

Data Download Date/Time: 05/29/2018 17:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANE,MARC,T
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: LANE,MARC
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 04/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 05/06/2018 00:09:58

Diagnostics Check	OK	00:03
Air Blank	0.000	00:03
Control Test	0.077	00:04
Air Blank	0.000	00:04
Subject Sample # 1	0.166	00:05
Breath Volume	2.031	00:05
Air Blank	0.000	00:06
Air Blank	0.000	00:07
Subject Sample # 2	0.162	00:08
Breath Volume	1.585	00:08
Air Blank	0.000	00:09
Control Test	0.076	00:09
Air Blank	0.000	00:09
Diagnostics Check	OK	00:09

Data Download Date/Time: 05/29/2018 17:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Port Richey PD
ARRESTING OFFICER: WILLIAMS,ROBERT
ARRESTING OFFICER AGENCY: Port Richey PD
LAST AGENCY INSPECTION: 04/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: JENK

DATE/TIME OF TEST: 05/06/2018 03:48:21

Diagnostics Check	OK	03:41
Air Blank	0.000	03:42
Control Test	0.077	03:42
Air Blank	0.000	03:43
Subject Sample # 1	0.220	03:43
Breath Volume	3.953	03:43
Air Blank	0.000	03:44
Air Blank	0.000	03:46
Subject Sample # 2	0.213	03:46
Breath Volume	3.773	03:46
Air Blank	0.000	03:47
Control Test	0.077	03:47
Air Blank	0.000	03:48
Diagnostics Check	OK	03:48

Data Download Date/Time: 05/29/2018 17:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,BARRY,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: NIXON,BARRY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 04/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 05/11/2018 01:01:14

Diagnostics Check	OK	00:55
Air Blank	0.000	00:55
Control Test	0.078	00:56
Air Blank	0.000	00:56
Subject Sample # 1	0.000	00:57
Breath Volume	2.632	00:57
Air Blank	0.000	00:57
Air Blank	0.000	00:59
Subject Sample # 2	0.000	00:59
Breath Volume	2.632	00:59
Air Blank	0.000	01:00
Control Test	0.078	01:00
Air Blank	0.000	01:01
Diagnostics Check	OK	01:01

Data Download Date/Time: 05/29/2018 17:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,BARRY,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: NIXON,BARRY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 04/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 05/18/2018 04:15:38

Diagnostics Check	OK	04:09
Air Blank	0.000	04:10
Control Test	0.077	04:10
Air Blank	0.000	04:10
Subject Sample # 1	0.129	04:11
Breath Volume	2.136	04:11
Air Blank	0.000	04:11
Air Blank	0.000	04:13
Subject Sample # 2	0.122	04:14
Breath Volume	1.863	04:14
Air Blank	0.000	04:14
Control Test	0.078	04:15
Air Blank	0.000	04:15
Diagnostics Check	OK	04:15

Data Download Date/Time: 05/29/2018 17:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,BARRY,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: NIXON,BARRY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 04/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: MARKS

DATE/TIME OF TEST: 05/19/2018 02:03:12

Diagnostics Check	OK	01:57
Air Blank	0.000	01:57
Control Test	0.078	01:58
Air Blank	0.000	01:58
Subject Sample # 1	0.000	01:59
Breath Volume	1.664	01:59
Air Blank	0.000	01:59
Air Blank	0.000	02:01
Subject Sample # 2	0.000	02:01
Breath Volume	1.453	02:01
Air Blank	0.000	02:02
Control Test	0.078	02:02
Air Blank	0.000	02:03
Diagnostics Check	OK	02:03

Data Download Date/Time: 05/29/2018 17:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVARADO,ANGEL,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ALVARADO,ANGEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: MULVANEY

DATE/TIME OF TEST: 05/19/2018 05:12:37

Diagnostics Check	OK	05:08	
Air Blank	0.000	05:09	
Control Test	0.078	05:09	
Air Blank	0.000	05:09	
Subject Sample	# 1	05:11	Subject Test Refused
Breath Volume	0.000	05:11	
Air Blank	0.000	05:11	
Control Test	0.078	05:12	
Air Blank	0.000	05:12	
Diagnostics Check	OK	05:12	

Data Download Date/Time: 05/29/2018 17:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAEZ HERNANDEZ,DENYS,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: PAEZ HERNANDEZ,DENYS
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:55

SUBJECT'S LAST NAME: CANNON

DATE/TIME OF TEST: 05/23/2018 07:26:48

Diagnostics Check	OK	07:16	
Air Blank	0.000	07:17	
Control Test	0.078	07:17	
Air Blank	0.000	07:17	
Subject Sample	# 1	07:20	No Sample Provided
Breath Volume	0.000	07:20	
Air Blank	0.000	07:21	
Air Blank	0.000	07:23	
Subject Sample	# 2	07:25	Subject Test Refused
Breath Volume	0.000	07:25	
Air Blank	0.000	07:25	
Control Test	0.077	07:26	
Air Blank	0.000	07:26	
Diagnostics Check	OK	07:26	

Data Download Date/Time: 05/29/2018 17:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,BARRY,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: NIXON,BARRY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 04/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: ANZALDO

DATE/TIME OF TEST: 05/23/2018 23:07:15

Diagnostics Check	OK	23:01
Air Blank	0.000	23:01
Control Test	0.078	23:02
Air Blank	0.000	23:02
Subject Sample # 1	0.000	23:02
Breath Volume	1.644	23:02
Air Blank	0.000	23:03
Air Blank	0.000	23:05
Subject Sample # 2	0.000	23:05
Breath Volume	1.667	23:05
Air Blank	0.000	23:06
Control Test	0.079	23:06
Air Blank	0.000	23:07
Diagnostics Check	OK	23:07

Data Download Date/Time: 05/29/2018 17:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COFFEE,JASON,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: COFFEE,JASON
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: DARLING

DATE/TIME OF TEST: 05/26/2018 22:09:11

Diagnostics Check	OK	22:02
Air Blank	0.000	22:03
Control Test	0.078	22:03
Air Blank	0.000	22:04
Subject Sample # 1	0.158	22:04
Breath Volume	2.457	22:04
Air Blank	0.000	22:05
Air Blank	0.000	22:07
Subject Sample # 2	0.159	22:07
Breath Volume	2.664	22:07
Air Blank	0.000	22:08
Control Test	0.077	22:08
Air Blank	0.000	22:09
Diagnostics Check	OK	22:09

Data Download Date/Time: 05/29/2018 17:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,BARRY,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: NIXON,BARRY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 04/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: THIEL

DATE/TIME OF TEST: 05/27/2018 03:02:43

Diagnostics Check	OK	02:56
Air Blank	0.000	02:56
Control Test	0.078	02:57
Air Blank	0.000	02:57
Subject Sample # 1	0.174	02:58
Breath Volume	2.863	02:58
Air Blank	0.000	02:58
Air Blank	0.000	03:00
Subject Sample # 2	0.173	03:01
Breath Volume	2.484	03:01
Air Blank	0.000	03:01
Control Test	0.078	03:02
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

Data Download Date/Time: 05/29/2018 17:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCE, SHNAYDINE, V
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SPENCE, SHNAYDINE
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 11:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: LA BARGE

DATE/TIME OF TEST: 05/28/2018 02:22:14

Diagnostics Check	OK	02:19	
Air Blank	0.000	02:19	
Control Test	0.077	02:20	
Air Blank	0.000	02:20	
Subject Sample	# 1	02:20	Subject Test Refused
Breath Volume	0.000	02:20	
Air Blank	0.000	02:21	
Control Test	0.077	02:21	
Air Blank	0.000	02:22	
Diagnostics Check	OK	02:22	

Data Download Date/Time: 05/29/2018 17:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVARADO,ANGEL,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ALVARADO,ANGEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:06

SUBJECT'S LAST NAME: MATHIS

DATE/TIME OF TEST: 05/28/2018 22:33:18

Diagnostics Check	OK	22:27
Air Blank	0.000	22:27
Control Test	0.076	22:27
Air Blank	0.000	22:28
Subject Sample # 1	0.243	22:28
Breath Volume	2.601	22:28
Air Blank	0.000	22:29
Air Blank	0.000	22:31
Subject Sample # 2	0.235	22:31
Breath Volume	2.515	22:31
Air Blank	0.000	22:32
Control Test	0.076	22:32
Air Blank	0.000	22:33
Diagnostics Check	OK	22:33

Data Download Date/Time: 05/29/2018 17:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIMPKINS,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: CLARK COTTO

DATE/TIME OF TEST: 05/02/2018 01:20:44

Diagnostics Check	OK	01:13
Air Blank	0.000	01:14
Control Test	0.079	01:14
Air Blank	0.000	01:15
Subject Sample # 1	0.000	01:16
Breath Volume	2.492	01:16
Air Blank	0.000	01:16
Air Blank	0.000	01:18
Subject Sample # 2	0.000	01:19
Breath Volume	2.378	01:19
Air Blank	0.000	01:19
Control Test	0.080	01:20
Air Blank	0.000	01:20
Diagnostics Check	OK	01:20

Data Download Date/Time: 05/03/2018 18:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,KAMIKA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MEDINA,P
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 05/02/2018 02:51:21

Diagnostics Check	OK	02:47	
Air Blank	0.000	02:48	
Control Test	0.079	02:48	
Air Blank	0.000	02:48	
Subject Sample	# 1	02:49	Subject Test Refused
Breath Volume	0.000	02:49	
Air Blank	0.000	02:50	
Control Test	0.079	02:50	
Air Blank	0.000	02:51	
Diagnostics Check	OK	02:51	

Data Download Date/Time: 05/03/2018 18:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HONEYGHON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: PATTERSON JR

DATE/TIME OF TEST: 05/02/2018 03:34:55

Diagnostics Check	OK	03:28
Air Blank	0.000	03:29
Control Test	0.080	03:29
Air Blank	0.000	03:30
Subject Sample # 1	0.000	03:30
Breath Volume	1.703	03:30
Air Blank	0.000	03:30
Air Blank	0.000	03:32
Subject Sample # 2	0.000	03:33
Breath Volume	2.105	03:33
Air Blank	0.000	03:34
Control Test	0.080	03:34
Air Blank	0.000	03:34
Diagnostics Check	OK	03:34

Data Download Date/Time: 05/03/2018 18:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MOREJON,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: BACON

DATE/TIME OF TEST: 05/03/2018 01:00:55

Diagnostics Check	OK	00:52
Air Blank	0.000	00:52
Control Test	0.081	00:53
Air Blank	0.000	00:53
Subject Sample # 1	0.110	00:55
Breath Volume	1.257	00:55
Air Blank	0.000	00:56
Air Blank	0.000	00:58
Subject Sample # 2	0.113	00:59
Breath Volume	1.308	00:59
Air Blank	0.000	01:00
Control Test	0.079	01:00
Air Blank	0.000	01:00
Diagnostics Check	OK	01:00

Data Download Date/Time: 05/03/2018 18:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,KAMIKA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIERRA,P
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 05/03/2018 01:17:32

Diagnostics Check	OK	01:08
Air Blank	0.000	01:09
Control Test	0.079	01:09
Air Blank	0.000	01:10
Subject Sample # 1	0.043	01:12
Breath Volume	2.558	01:12
Air Blank	0.000	01:13
Air Blank	0.000	01:15
Subject Sample # 2	0.036	01:16
Breath Volume	1.304	01:16
Air Blank	0.000	01:16
Control Test	0.080	01:16
Air Blank	0.000	01:17
Diagnostics Check	OK	01:17

Data Download Date/Time: 05/03/2018 18:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKIPPER,ROGER,G
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MUNEZ,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: OTHER
ARREST TIME: 13:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:30

SUBJECT'S LAST NAME: FARMER

DATE/TIME OF TEST: 05/03/2018 14:10:24

Diagnostics Check	OK	14:07
Air Blank	0.000	14:07
Control Test	0.081	14:07
Air Blank	0.000	14:08
Subject Sample # 1	0.000	14:08
Breath Volume	3.062	14:08
Air Blank	0.000	14:09
Air Blank	0.000	14:10

Data Download Date/Time: 05/03/2018 18:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,KAMIKA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HONEYGHON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: VELASQUEZ

DATE/TIME OF TEST: 05/05/2018 22:51:06

Diagnostics Check	OK	22:44
Air Blank	0.000	22:44
Control Test	0.081	22:45
Air Blank	0.000	22:45
Subject Sample # 1	0.173	22:46
Breath Volume	1.144	22:46
Air Blank	0.000	22:46
Air Blank	0.000	22:48
Subject Sample # 2	0.181	22:49
Breath Volume	1.863	22:49
Air Blank	0.000	22:50
Control Test	0.078	22:50
Air Blank	0.000	22:51
Diagnostics Check	OK	22:51

Data Download Date/Time: 06/19/2018 18:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WOLSTENHOLM,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: PONGRATZ

DATE/TIME OF TEST: 05/16/2018 23:57:51

Diagnostics Check	OK	23:51
Air Blank	0.000	23:52
Control Test	0.080	23:52
Air Blank	0.000	23:52
Subject Sample # 1	0.129	23:53
Breath Volume	2.441	23:53
Air Blank	0.000	23:54
Air Blank	0.000	23:55
Subject Sample # 2	0.125	23:56
Breath Volume	1.656	23:56
Air Blank	0.000	23:56
Control Test	0.079	23:57
Air Blank	0.000	23:57
Diagnostics Check	OK	23:57

Data Download Date/Time: 06/19/2018 18:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRANKLIN,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: DRAWDY

DATE/TIME OF TEST: 05/17/2018 03:06:38

Diagnostics Check	OK	03:00
Air Blank	0.000	03:00
Control Test	0.081	03:01
Air Blank	0.000	03:01
Subject Sample # 1	0.136	03:02
Breath Volume	1.828	03:02
Air Blank	0.000	03:02
Air Blank	0.000	03:04
Subject Sample # 2	0.135	03:05
Breath Volume	1.789	03:05
Air Blank	0.000	03:05
Control Test	0.079	03:06
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

Data Download Date/Time: 06/19/2018 18:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKIPPER,ROGER,G
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WORKMAN,R 4332
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 08:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:47

SUBJECT'S LAST NAME: BENCZI JR

DATE/TIME OF TEST: 05/17/2018 10:14:49

Diagnostics Check	OK	10:08
Air Blank	0.000	10:08
Control Test	0.080	10:09
Air Blank	0.000	10:09
Subject Sample # 1	0.000	10:10
Breath Volume	2.640	10:10
Air Blank	0.000	10:10
Air Blank	0.000	10:12
Subject Sample # 2	0.000	10:13
Breath Volume	1.917	10:13
Air Blank	0.000	10:13
Control Test	0.081	10:14
Air Blank	0.000	10:14
Diagnostics Check	OK	10:14

Data Download Date/Time: 06/19/2018 18:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: EVANS,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: ALGIER PANCIERA

DATE/TIME OF TEST: 05/17/2018 22:05:42

Diagnostics Check	OK	21:57	
Air Blank	0.000	21:57	
Control Test	0.081	21:58	
Air Blank	0.000	21:58	
Subject Sample	# 1	22:01	No Sample Provided
Breath Volume	0.000	22:01	
Air Blank	0.000	22:02	
Air Blank	0.000	22:04	
Subject Sample	# 2	22:04	Subject Test Refused
Breath Volume	0.000	22:04	
Air Blank	0.000	22:04	
Control Test	0.080	22:05	
Air Blank	0.000	22:05	
Diagnostics Check	OK	22:05	

Data Download Date/Time: 06/19/2018 18:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,KAMIKA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANGELICCHI,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 06:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:21

SUBJECT'S LAST NAME: SUAREZ

DATE/TIME OF TEST: 05/18/2018 06:48:09

Diagnostics Check	OK	06:41
Air Blank	0.000	06:42
Control Test	0.080	06:42
Air Blank	0.000	06:43
Subject Sample # 1	0.099	06:43
Breath Volume	2.652	06:43
Air Blank	0.000	06:44
Air Blank	0.000	06:46
Subject Sample # 2	0.098	06:46
Breath Volume	2.972	06:46
Air Blank	0.000	06:47
Control Test	0.080	06:47
Air Blank	0.000	06:48
Diagnostics Check	OK	06:48

Data Download Date/Time: 06/19/2018 18:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKIPPER,ROGER,G
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HOSKINS,H 4425
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 10:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:32

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 05/18/2018 12:01:09

Diagnostics Check	OK	11:53
Air Blank	0.000	11:54
Control Test	0.080	11:54
Air Blank	0.000	11:55
Subject Sample # 1	0.107	11:55
Breath Volume	1.664	11:55
Air Blank	0.000	11:56
Air Blank	0.000	11:58
Subject Sample # 2	0.108	11:59
Breath Volume	1.710	11:59
Air Blank	0.000	12:00
Control Test	0.081	12:00
Air Blank	0.000	12:01
Diagnostics Check	OK	12:01

Data Download Date/Time: 06/19/2018 18:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,KAMIKA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MOREJON,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:44

SUBJECT'S LAST NAME: PEREZ VAZQUEZ

DATE/TIME OF TEST: 05/18/2018 21:11:45

Diagnostics Check	OK	21:05
Air Blank	0.000	21:05
Control Test	0.081	21:05
Air Blank	0.000	21:06
Subject Sample # 1	0.218	21:07
Breath Volume	2.429	21:07
Air Blank	0.000	21:07
Air Blank	0.000	21:09
Subject Sample # 2	0.214	21:10
Breath Volume	3.023	21:10
Air Blank	0.000	21:10
Control Test	0.077	21:11
Air Blank	0.000	21:11
Diagnostics Check	OK	21:11

Data Download Date/Time: 06/19/2018 18:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS, RONALD, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: LOPEZ-HERNANDEZ

DATE/TIME OF TEST: 05/19/2018 01:55:53

Diagnostics Check	OK	01:49
Air Blank	0.000	01:50
Control Test	0.080	01:50
Air Blank	0.000	01:51
Subject Sample # 1	0.238	01:51
Breath Volume	1.363	01:51
Air Blank	0.000	01:52
Air Blank	0.000	01:53
Subject Sample # 2	0.245	01:54
Breath Volume	1.992	01:54
Air Blank	0.000	01:54
Control Test	0.079	01:55
Air Blank	0.000	01:55
Diagnostics Check	OK	01:55

Data Download Date/Time: 06/19/2018 18:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIMPKINS, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: ROCHE

DATE/TIME OF TEST: 05/19/2018 04:01:40

Diagnostics Check	OK	03:50	
Air Blank	0.000	03:50	
Control Test	0.080	03:51	
Air Blank	0.000	03:51	
Subject Sample	# 1	03:54	No Sample Provided
Breath Volume	0.000	03:54	
Air Blank	0.000	03:55	
Air Blank	0.000	03:57	
Subject Sample	# 2	04:00	No Sample Provided
Breath Volume	0.000	04:00	
Air Blank	0.000	04:00	
Control Test	0.080	04:01	
Air Blank	0.000	04:01	
Diagnostics Check	OK	04:01	

Data Download Date/Time: 06/19/2018 18:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DRISCOLL,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:02

SUBJECT'S LAST NAME: PALMER

DATE/TIME OF TEST: 05/19/2018 06:30:32

Diagnostics Check	OK	06:23
Air Blank	0.000	06:23
Control Test	0.080	06:24
Air Blank	0.000	06:24
Subject Sample # 1	0.145	06:25
Breath Volume	1.648	06:25
Air Blank	0.000	06:26
Air Blank	0.000	06:28
Subject Sample # 2	0.139	06:28
Breath Volume	1.378	06:28
Air Blank	0.000	06:29
Control Test	0.079	06:29
Air Blank	0.000	06:30
Diagnostics Check	OK	06:30

Data Download Date/Time: 06/19/2018 18:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: EDGE,S
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: SOTELO-POPOCA

DATE/TIME OF TEST: 05/19/2018 23:15:30

Diagnostics Check	OK	23:08
Air Blank	0.000	23:08
Control Test	0.080	23:08
Air Blank	0.000	23:09
Subject Sample # 1	0.122	23:10
Breath Volume	3.886	23:10
Air Blank	0.000	23:11
Air Blank	0.000	23:13
Subject Sample # 2	0.119	23:13
Breath Volume	3.921	23:13
Air Blank	0.000	23:14
Control Test	0.080	23:14
Air Blank	0.000	23:15
Diagnostics Check	OK	23:15

Data Download Date/Time: 06/19/2018 18:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIERRA,P
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 05/20/2018 00:48:07

Diagnostics Check	OK	00:41
Air Blank	0.000	00:41
Control Test	0.080	00:42
Air Blank	0.000	00:42
Subject Sample # 1	0.169	00:43
Breath Volume	1.648	00:43
Air Blank	0.000	00:43
Air Blank	0.000	00:45
Subject Sample # 2	0.163	00:46
Breath Volume	1.554	00:46
Air Blank	0.000	00:47
Control Test	0.079	00:47
Air Blank	0.000	00:48
Diagnostics Check	OK	00:48

Data Download Date/Time: 06/19/2018 18:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MEDINA,P
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: JUAREZ

DATE/TIME OF TEST: 05/20/2018 01:33:32

Diagnostics Check	OK	01:24
Air Blank	0.000	01:24
Control Test	0.079	01:24
Air Blank	0.000	01:25
Subject Sample # 1	0.173	01:26
Breath Volume	1.457	01:26
Air Blank	0.000	01:26
Air Blank	0.000	01:28
Subject Sample # 2	0.196	01:29
Breath Volume	2.992	01:29
Air Blank	0.000	01:29
Air Blank	0.000	01:31
Subject Sample # 3	0.180	01:31
Breath Volume	1.535	01:31
Air Blank	0.000	01:32
Control Test	0.078	01:32
Air Blank	0.000	01:33
Diagnostics Check	OK	01:33

Data Download Date/Time: 06/19/2018 18:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NIEUWENDAM,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: SUAREZ-HERNANDEZ

DATE/TIME OF TEST: 05/20/2018 02:09:26

Diagnostics Check	OK	02:02
Air Blank	0.000	02:03
Control Test	0.079	02:03
Air Blank	0.000	02:04
Subject Sample # 1	0.212	02:04
Breath Volume	2.175	02:04
Air Blank	0.000	02:05
Air Blank	0.000	02:07
Subject Sample # 2	0.208	02:07
Breath Volume	2.156	02:07
Air Blank	0.000	02:08
Control Test	0.078	02:08
Air Blank	0.000	02:09
Diagnostics Check	OK	02:09

Data Download Date/Time: 06/19/2018 18:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: KINSEY

DATE/TIME OF TEST: 05/20/2018 03:45:01

Diagnostics Check	OK	03:38
Air Blank	0.000	03:39
Control Test	0.079	03:39
Air Blank	0.000	03:39
Subject Sample # 1	0.000	03:40
Breath Volume	1.652	03:40
Air Blank	0.000	03:41
Air Blank	0.000	03:43
Subject Sample # 2	0.000	03:43
Breath Volume	1.847	03:43
Air Blank	0.000	03:44
Control Test	0.079	03:44
Air Blank	0.000	03:44
Diagnostics Check	OK	03:45

Data Download Date/Time: 06/19/2018 18:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: CAX-GARCIA

DATE/TIME OF TEST: 05/20/2018 04:23:34

Diagnostics Check	OK	04:16
Air Blank	0.000	04:17
Control Test	0.080	04:17
Air Blank	0.000	04:18
Subject Sample # 1	0.148	04:18
Breath Volume	2.417	04:18
Air Blank	0.000	04:19
Air Blank	0.000	04:21
Subject Sample # 2	0.149	04:21
Breath Volume	2.468	04:21
Air Blank	0.000	04:22
Control Test	0.079	04:22
Air Blank	0.000	04:23
Diagnostics Check	OK	04:23

Data Download Date/Time: 06/19/2018 18:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SANTOS,J
ARRESTING OFFICER AGENCY: Macdill Air Force Ba
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 07:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:13

SUBJECT'S LAST NAME: KELLER

DATE/TIME OF TEST: 05/20/2018 08:41:25

Diagnostics Check	OK	08:34
Air Blank	0.000	08:35
Control Test	0.080	08:35
Air Blank	0.000	08:36
Subject Sample # 1	0.136	08:36
Breath Volume	1.855	08:36
Air Blank	0.000	08:37
Air Blank	0.000	08:39
Subject Sample # 2	0.140	08:39
Breath Volume	1.867	08:39
Air Blank	0.000	08:40
Control Test	0.079	08:40
Air Blank	0.000	08:41
Diagnostics Check	OK	08:41

Data Download Date/Time: 06/19/2018 18:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,REBECCA,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: O'DONNELL,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: O'BRIEN

DATE/TIME OF TEST: 05/20/2018 18:29:18

Diagnostics Check	OK	18:22
Air Blank	0.000	18:23
Control Test	0.080	18:23
Air Blank	0.000	18:24
Subject Sample # 1	0.167	18:24
Breath Volume	2.515	18:24
Air Blank	0.000	18:25
Air Blank	0.000	18:27
Subject Sample # 2	0.164	18:27
Breath Volume	2.554	18:27
Air Blank	0.000	18:28
Control Test	0.079	18:28
Air Blank	0.000	18:29
Diagnostics Check	OK	18:29

Data Download Date/Time: 06/19/2018 18:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MOREJON,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:08

SUBJECT'S LAST NAME: MORA-VIVEROS

DATE/TIME OF TEST: 05/21/2018 05:36:53

Diagnostics Check	OK	05:29
Air Blank	0.000	05:30
Control Test	0.081	05:30
Air Blank	0.000	05:31
Subject Sample # 1	0.160	05:32
Breath Volume	1.964	05:32
Air Blank	0.000	05:33
Air Blank	0.000	05:34
Subject Sample # 2	0.157	05:35
Breath Volume	1.492	05:35
Air Blank	0.000	05:35
Control Test	0.079	05:36
Air Blank	0.000	05:36
Diagnostics Check	OK	05:36

Data Download Date/Time: 06/19/2018 18:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: FAVORS

DATE/TIME OF TEST: 05/21/2018 23:47:12

Diagnostics Check	OK	23:40
Air Blank	0.000	23:40
Control Test	0.080	23:41
Air Blank	0.000	23:41
Subject Sample # 1	0.170	23:42
Breath Volume	2.042	23:42
Air Blank	0.000	23:43
Air Blank	0.000	23:44
Subject Sample # 2	0.170	23:45
Breath Volume	2.023	23:45
Air Blank	0.000	23:46
Control Test	0.079	23:46
Air Blank	0.000	23:47
Diagnostics Check	OK	23:47

Data Download Date/Time: 06/19/2018 18:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS, RONALD, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER, A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: SAPIENZA

DATE/TIME OF TEST: 05/22/2018 21:37:04

Diagnostics Check	OK	21:30
Air Blank	0.000	21:31
Control Test	0.080	21:31
Air Blank	0.000	21:32
Subject Sample # 1	0.138	21:32
Breath Volume	3.578	21:32
Air Blank	0.000	21:33
Air Blank	0.000	21:35
Subject Sample # 2	0.136	21:35
Breath Volume	3.484	21:35
Air Blank	0.000	21:36
Control Test	0.079	21:36
Air Blank	0.000	21:36
Diagnostics Check	OK	21:37

Data Download Date/Time: 06/19/2018 18:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,KAMIKA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: STEERE

DATE/TIME OF TEST: 05/23/2018 00:55:07

Diagnostics Check	OK	00:48
Air Blank	0.000	00:49
Control Test	0.080	00:49
Air Blank	0.000	00:50
Subject Sample # 1	0.092	00:50
Breath Volume	2.375	00:50
Air Blank	0.000	00:51
Air Blank	0.000	00:53
Subject Sample # 2	0.089	00:53
Breath Volume	2.546	00:53
Air Blank	0.000	00:54
Control Test	0.079	00:54
Air Blank	0.000	00:55
Diagnostics Check	OK	00:55

Data Download Date/Time: 06/19/2018 18:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LABIB, A
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 05/23/2018 02:23:03

Diagnostics Check	OK	02:16
Air Blank	0.000	02:16
Control Test	0.080	02:16
Air Blank	0.000	02:17
Subject Sample # 1	0.128	02:18
Breath Volume	1.261	02:18
Air Blank	0.000	02:18
Air Blank	0.000	02:20
Subject Sample # 2	0.123	02:21
Breath Volume	1.933	02:21
Air Blank	0.000	02:22
Control Test	0.079	02:22
Air Blank	0.000	02:22
Diagnostics Check	OK	02:23

Data Download Date/Time: 06/19/2018 18:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003406

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHIVER,JAMES,W
BREATH TEST OPERATOR AGENCY: Liberty County SO
ARRESTING OFFICER: SHIVER,JAMES
ARRESTING OFFICER AGENCY: Liberty County SO
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: GIRON

DATE/TIME OF TEST: 05/12/2018 21:30:26

Diagnostics Check	OK	21:21
Air Blank	0.000	21:21
Control Test	0.082	21:22
Air Blank	0.000	21:22
Subject Sample # 1	0.385	21:23
Breath Volume	1.121	21:23
Air Blank	0.000	21:23
Air Blank	0.000	21:25
Subject Sample # 2	0.364	21:25
Breath Volume	1.886	21:25
Air Blank	0.000	21:26
Air Blank	0.000	21:28
Subject Sample # 3	0.399	21:28
Breath Volume	2.800	21:28
Air Blank	0.000	21:29
Control Test	0.081	21:29
Air Blank	0.000	21:30
Diagnostics Check	OK	21:30

Data Download Date/Time: 05/16/2018 09:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERO,JOSE,I
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: HUGGINS,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: ALEXANDER

DATE/TIME OF TEST: 05/05/2018 03:46:04

Diagnostics Check	OK	03:39
Air Blank	0.000	03:40
Control Test	0.081	03:40
Air Blank	0.000	03:41
Subject Sample # 1	0.172	03:41
Breath Volume	2.785	03:41
Air Blank	0.000	03:42
Air Blank	0.000	03:44
Subject Sample # 2	0.172	03:44
Breath Volume	2.355	03:44
Air Blank	0.000	03:45
Control Test	0.081	03:45
Air Blank	0.000	03:45
Diagnostics Check	OK	03:46

Data Download Date/Time: 05/29/2018 11:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REGIS,STEVE,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: REGIS,STEVE
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: TALLEYRAND

DATE/TIME OF TEST: 05/11/2018 04:54:28

Diagnostics Check	OK	04:48
Air Blank	0.000	04:48
Control Test	0.081	04:48
Air Blank	0.000	04:49
Subject Sample # 1	0.238	04:49
Breath Volume	2.769	04:49
Air Blank	0.000	04:50
Air Blank	0.000	04:52
Subject Sample # 2	0.231	04:52
Breath Volume	2.773	04:52
Air Blank	0.000	04:53
Control Test	0.080	04:53
Air Blank	0.000	04:54
Diagnostics Check	OK	04:54

Data Download Date/Time: 05/29/2018 11:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERO,JOSE,I
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: RIVERO,JOSE
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: ALCANTARA

DATE/TIME OF TEST: 05/13/2018 00:46:55

Diagnostics Check	OK	00:44	
Air Blank	0.000	00:44	
Control Test	0.082	00:44	
Air Blank	0.000	00:45	
Subject Sample	# 1	00:45	Subject Test Refused
Breath Volume	0.000	00:45	
Air Blank	0.000	00:46	
Control Test	0.082	00:46	
Air Blank	0.000	00:46	
Diagnostics Check	OK	00:46	

Data Download Date/Time: 05/29/2018 11:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARRERO,JUAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: VINCENT,CARL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 12:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: ANTEZANA

DATE/TIME OF TEST: 05/18/2018 02:10:00

Diagnostics Check	OK	02:03
Air Blank	0.000	02:04
Control Test	0.082	02:04
Air Blank	0.000	02:05
Subject Sample # 1	0.099	02:05
Breath Volume	1.898	02:05
Air Blank	0.000	02:06
Air Blank	0.000	02:07
Subject Sample # 2	0.104	02:08
Breath Volume	1.683	02:08
Air Blank	0.000	02:09
Control Test	0.081	02:09
Air Blank	0.000	02:09
Diagnostics Check	OK	02:10

Data Download Date/Time: 05/29/2018 11:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAVINON,CARLOS,M
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: SAVINON,CARLOS
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 19:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: VALENTIN

DATE/TIME OF TEST: 05/19/2018 20:55:33

Diagnostics Check	OK	20:48
Air Blank	0.000	20:48
Control Test	0.082	20:49
Air Blank	0.000	20:49
Subject Sample # 1	0.115	20:51
Breath Volume	1.531	20:51
Air Blank	0.000	20:51
Air Blank	0.000	20:53
Subject Sample # 2	0.104	20:54
Breath Volume	1.382	20:54
Air Blank	0.000	20:54
Control Test	0.081	20:54
Air Blank	0.000	20:55
Diagnostics Check	OK	20:55

Data Download Date/Time: 05/29/2018 11:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LA ROSA CASTRO,DAVID,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: LA ROSA CASTRO,DAVID
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: PENA

DATE/TIME OF TEST: 05/23/2018 04:12:39

Diagnostics Check	OK	04:06
Air Blank	0.000	04:06
Control Test	0.082	04:07
Air Blank	0.000	04:07
Subject Sample # 1	0.058	04:08
Breath Volume	3.082	04:08
Air Blank	0.000	04:08
Air Blank	0.000	04:10
Subject Sample # 2	0.057	04:11
Breath Volume	2.480	04:11
Air Blank	0.000	04:11
Control Test	0.082	04:12
Air Blank	0.000	04:12
Diagnostics Check	OK	04:12

Data Download Date/Time: 05/29/2018 11:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIERRA,ALIEK,O
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: SIERRA,ALIEK
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: ALMEIDA

DATE/TIME OF TEST: 05/26/2018 02:32:37

Diagnostics Check	OK	02:23	
Air Blank	0.000	02:24	
Control Test	0.081	02:24	
Air Blank	0.000	02:25	
Subject Sample	# 1	02:28	No Sample Provided
Breath Volume	0.000	02:28	
Air Blank	0.000	02:28	
Air Blank	0.000	02:30	
Subject Sample	# 2	02:31	Subject Test Refused
Breath Volume	0.000	02:31	
Air Blank	0.000	02:31	
Control Test	0.081	02:32	
Air Blank	0.000	02:32	
Diagnostics Check	OK	02:32	

Data Download Date/Time: 05/29/2018 11:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003410

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZIGLER,RICKIE,R
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: ZIGLER,RICKIE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 437738
DRY GAS EXPIRY DATE: 06/08/2019
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: GRISHAM

DATE/TIME OF TEST: 05/19/2018 00:07:56

Diagnostics Check	OK	00:00
Air Blank	0.000	00:01
Control Test	0.080	00:01
Air Blank	0.000	00:02
Subject Sample # 1	0.143	00:03
Breath Volume	1.597	00:03
Air Blank	0.000	00:03
Air Blank	0.000	00:05
Subject Sample # 2	0.138	00:06
Breath Volume	1.605	00:06
Air Blank	0.000	00:07
Control Test	0.081	00:07
Air Blank	0.000	00:07
Diagnostics Check	OK	00:07

Data Download Date/Time: 06/26/2018 18:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003410

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILDEBRANDT,PATRICK,A
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: RICKERSON,KENTON
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 437738
DRY GAS EXPIRY DATE: 06/08/2019
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: STAFFORD

DATE/TIME OF TEST: 05/24/2018 00:28:13

Diagnostics Check	OK	00:21
Air Blank	0.000	00:22
Control Test	0.080	00:22
Air Blank	0.000	00:23
Subject Sample # 1	0.151	00:23
Breath Volume	1.570	00:23
Air Blank	0.000	00:24
Air Blank	0.000	00:26
Subject Sample # 2	0.151	00:26
Breath Volume	1.511	00:26
Air Blank	0.000	00:27
Control Test	0.082	00:27
Air Blank	0.000	00:28
Diagnostics Check	OK	00:28

Data Download Date/Time: 06/26/2018 18:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003410

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZIGLER,RICKIE,R
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: ZIGLER,RICKIE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 437738
DRY GAS EXPIRY DATE: 06/08/2019
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: AULT

DATE/TIME OF TEST: 05/26/2018 23:44:44

Diagnostics Check	OK	23:41	
Air Blank	0.000	23:42	
Control Test	0.080	23:42	
Air Blank	0.000	23:43	
Subject Sample	# 1	23:43	Subject Test Refused
Breath Volume	0.000	23:43	
Air Blank	0.000	23:43	
Control Test	0.081	23:44	
Air Blank	0.000	23:44	
Diagnostics Check	OK	23:44	

Data Download Date/Time: 06/26/2018 18:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: CODY

DATE/TIME OF TEST: 05/05/2018 04:30:56

Diagnostics Check	OK	04:28	
Air Blank	0.000	04:29	
Control Test	0.080	04:29	
Air Blank	0.000	04:29	
Subject Sample	# 1	04:30	Subject Test Refused
Breath Volume	0.000	04:30	
Air Blank		04:30	RFI Detect
Air Blank	0.000	04:30	

Data Download Date/Time: 05/27/2018 17:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: TOLENTINO

DATE/TIME OF TEST: 05/06/2018 02:48:49

Diagnostics Check	OK	02:42
Air Blank	0.000	02:43
Control Test	0.080	02:43
Air Blank	0.000	02:44
Subject Sample # 1	0.175	02:44
Breath Volume	2.304	02:44
Air Blank	0.000	02:45
Air Blank	0.000	02:46
Subject Sample # 2	0.176	02:47
Breath Volume	2.640	02:47
Air Blank	0.000	02:47
Control Test	0.079	02:48
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

Data Download Date/Time: 05/27/2018 17:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: ATWOOD

DATE/TIME OF TEST: 05/10/2018 22:52:56

Diagnostics Check	OK	22:46
Air Blank	0.000	22:47
Control Test	0.080	22:47
Air Blank	0.000	22:47
Subject Sample # 1	0.044	22:48
Breath Volume	1.179	22:48
Air Blank	0.000	22:49
Air Blank	0.000	22:50
Subject Sample # 2	0.045	22:51
Breath Volume	1.515	22:51
Air Blank	0.000	22:52
Control Test	0.080	22:52
Air Blank	0.000	22:52
Diagnostics Check	OK	22:52

Data Download Date/Time: 05/27/2018 17:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: BALTUSNIK

DATE/TIME OF TEST: 05/25/2018 01:49:07

Diagnostics Check	OK	01:42
Air Blank	0.000	01:43
Control Test	0.079	01:43
Air Blank	0.000	01:44
Subject Sample # 1	0.024	01:44
Breath Volume	1.949	01:44
Air Blank	0.000	01:45
Air Blank	0.000	01:47
Subject Sample # 2	0.019	01:47
Breath Volume	1.562	01:47
Air Blank	0.000	01:48
Control Test	0.079	01:48
Air Blank	0.000	01:49
Diagnostics Check	OK	01:49

Data Download Date/Time: 05/27/2018 17:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: CHAVES LOPEZ

DATE/TIME OF TEST: 05/27/2018 02:54:42

Diagnostics Check	OK	02:48
Air Blank	0.000	02:48
Control Test	0.080	02:49
Air Blank	0.000	02:49
Subject Sample # 1	0.121	02:50
Breath Volume	1.253	02:50
Air Blank	0.000	02:50
Air Blank	0.000	02:52
Subject Sample # 2	0.125	02:53
Breath Volume	1.566	02:53
Air Blank	0.000	02:53
Control Test	0.079	02:54
Air Blank	0.000	02:54
Diagnostics Check	OK	02:54

Data Download Date/Time: 05/27/2018 17:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPOS, MICHAEL, A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CAMPOS, MICHAEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: DALTON

DATE/TIME OF TEST: 05/11/2018 23:01:10

Diagnostics Check	OK	22:55
Air Blank	0.000	22:55
Control Test	0.078	22:55
Air Blank	0.000	22:56
Subject Sample # 1	0.169	22:56
Breath Volume	3.492	22:56
Air Blank	0.000	22:57
Air Blank	0.000	22:59
Subject Sample # 2	0.165	22:59
Breath Volume	3.664	22:59
Air Blank	0.000	23:00
Control Test	0.077	23:00
Air Blank	0.000	23:01
Diagnostics Check	OK	23:01

Data Download Date/Time: 06/05/2018 22:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEEKMAN,JACOB,C
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: BEEKMAN,JACOB
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 03:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: SOVRAN

DATE/TIME OF TEST: 05/12/2018 04:22:12

Diagnostics Check	OK	04:08	
Air Blank	0.000	04:09	
Control Test	0.077	04:09	
Air Blank	0.000	04:09	
Subject Sample	# 1	04:13	No Sample Provided
Breath Volume	0.000	04:13	
Air Blank	0.000	04:13	
Air Blank	0.000	04:15	
Subject Sample	# 2	0.137	04:16
Breath Volume	1.789	04:16	
Air Blank	0.000	04:17	
Air Blank	0.000	04:19	
Subject Sample	# 3	0.135	04:20
Breath Volume	2.621	04:20	
Air Blank	0.000	04:21	
Control Test	0.078	04:21	
Air Blank	0.000	04:22	
Diagnostics Check	OK	04:22	

Data Download Date/Time: 06/05/2018 22:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPOS,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CAMPOS,MICHAEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 05/13/2018 03:46:16

Diagnostics Check	OK	03:40
Air Blank	0.000	03:40
Control Test	0.078	03:40
Air Blank	0.000	03:41
Subject Sample # 1	0.203	03:41
Breath Volume	2.285	03:41
Air Blank	0.000	03:42
Air Blank	0.000	03:44
Subject Sample # 2	0.199	03:44
Breath Volume	2.429	03:44
Air Blank	0.000	03:45
Control Test	0.077	03:45
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

Data Download Date/Time: 06/05/2018 22:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEEKMAN,JACOB,C
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: BEEKMAN,JACOB
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:08

SUBJECT'S LAST NAME: SMYK

DATE/TIME OF TEST: 05/13/2018 21:45:12

Diagnostics Check	OK	21:38
Air Blank	0.000	21:39
Control Test	0.078	21:39
Air Blank	0.000	21:39
Subject Sample # 1	0.261	21:40
Breath Volume	3.000	21:40
Air Blank	0.000	21:41
Air Blank	0.000	21:43
Subject Sample # 2	0.256	21:43
Breath Volume	2.863	21:43
Air Blank	0.000	21:44
Control Test	0.077	21:44
Air Blank	0.000	21:45
Diagnostics Check	OK	21:45

Data Download Date/Time: 06/05/2018 22:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPOS,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CAMPOS,MICHAEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: NAKONECZNYJ

DATE/TIME OF TEST: 05/30/2018 20:32:36

Diagnostics Check	OK	20:26
Air Blank	0.000	20:27
Control Test	0.077	20:27
Air Blank	0.000	20:27
Subject Sample # 1	0.223	20:28
Breath Volume	1.417	20:28
Air Blank	0.000	20:28
Air Blank	0.000	20:30
Subject Sample # 2	0.230	20:31
Breath Volume	1.902	20:31
Air Blank	0.000	20:31
Control Test	0.078	20:32
Air Blank	0.000	20:32
Diagnostics Check	OK	20:32

Data Download Date/Time: 06/05/2018 22:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTLE,CORY,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SUTTLE,CORY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: PEREZ-ALONSO

DATE/TIME OF TEST: 05/30/2018 22:37:51

Diagnostics Check	OK	22:31
Air Blank	0.000	22:32
Control Test	0.079	22:32
Air Blank	0.000	22:33
Subject Sample # 1	0.225	22:33
Breath Volume	2.457	22:33
Air Blank	0.000	22:34
Air Blank	0.000	22:35
Subject Sample # 2	0.227	22:36
Breath Volume	2.195	22:36
Air Blank	0.000	22:36
Control Test	0.079	22:37
Air Blank	0.000	22:37
Diagnostics Check	OK	22:37

Data Download Date/Time: 06/25/2018 10:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEEKMAN,JACOB,C
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: BEEKMAN,JACOB
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 05/13/2018 01:44:05

Diagnostics Check	OK	01:28	
Air Blank	0.000	01:29	
Control Test	0.079	01:29	
Air Blank	0.000	01:30	
Subject Sample	# 1 0.148	01:33	Volume Not Met
Breath Volume	0.171	01:33	
Air Blank	0.000	01:33	
Air Blank	0.000	01:35	
Subject Sample	# 2 0.200	01:38	No .020 agreement
Breath Volume	1.554	01:38	
Air Blank	0.000	01:38	
Air Blank	0.000	01:40	
Subject Sample	# 3 0.226	01:42	No .020 agreement
Breath Volume	1.574	01:42	
Air Blank	0.000	01:43	
Control Test	0.076	01:43	
Air Blank	0.000	01:43	
Diagnostics Check	OK	01:44	

Data Download Date/Time: 06/05/2018 22:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEEKMAN,JACOB,C
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: BEEKMAN,JACOB
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 05/13/2018 02:35:43

Diagnostics Check	OK	02:26
Air Blank	0.000	02:27
Control Test	0.077	02:27
Air Blank	0.000	02:28
Subject Sample # 1	0.196	02:29
Breath Volume	1.125	02:29
Air Blank	0.000	02:30
Air Blank	0.000	02:31
Subject Sample # 2	0.191	02:34
Breath Volume	1.355	02:34
Air Blank	0.000	02:34
Control Test	0.077	02:35
Air Blank	0.000	02:35
Diagnostics Check	OK	02:35

Data Download Date/Time: 06/05/2018 22:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILHELM, TONY, J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: STEVENS, MATTHEW
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 16:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:25

SUBJECT'S LAST NAME: MAYBERRY

DATE/TIME OF TEST: 05/03/2018 16:48:43

Diagnostics Check	OK	16:45	
Air Blank	0.000	16:46	
Control Test	0.080	16:46	
Air Blank	0.000	16:47	
Subject Sample	# 1	16:47	Subject Test Refused
Breath Volume	0.000	16:47	
Air Blank	0.000	16:47	
Control Test	0.080	16:48	
Air Blank	0.000	16:48	
Diagnostics Check	OK	16:48	

Data Download Date/Time: 06/06/2018 08:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITMAN,AARON,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: PITMAN,AARON
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: PELLMAN

DATE/TIME OF TEST: 05/06/2018 00:40:15

Diagnostics Check	OK	00:37	
Air Blank	0.000	00:37	
Control Test	0.079	00:38	
Air Blank	0.000	00:38	
Subject Sample	# 1	00:38	Subject Test Refused
Breath Volume	0.000	00:38	
Air Blank	0.000	00:39	
Control Test	0.080	00:39	
Air Blank	0.000	00:40	
Diagnostics Check	OK	00:40	

Data Download Date/Time: 06/06/2018 08:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,DAVID,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: HAMILTON,RAY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 15:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:25

SUBJECT'S LAST NAME: SAPPENFIELD

DATE/TIME OF TEST: 05/06/2018 16:56:58

Diagnostics Check	OK	16:53	
Air Blank	0.000	16:54	
Control Test	0.081	16:54	
Air Blank	0.000	16:55	
Subject Sample	# 1	16:55	Subject Test Refused
Breath Volume	0.000	16:55	
Air Blank	0.000	16:56	
Control Test	0.080	16:56	
Air Blank	0.000	16:56	
Diagnostics Check	OK	16:56	

Data Download Date/Time: 06/06/2018 08:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIMBRIL,JOSHUA,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ALEXANDER,C
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 19:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: LUEBBERT

DATE/TIME OF TEST: 05/11/2018 19:38:38

Diagnostics Check	OK	19:31
Air Blank	0.000	19:32
Control Test	0.080	19:32
Air Blank	0.000	19:33
Subject Sample # 1	0.045	19:33
Breath Volume	1.687	19:33
Air Blank	0.000	19:34
Air Blank	0.000	19:36
Subject Sample # 2	0.046	19:37
Breath Volume	1.949	19:37
Air Blank	0.000	19:37
Control Test	0.080	19:38
Air Blank	0.000	19:38
Diagnostics Check	OK	19:38

Data Download Date/Time: 06/06/2018 08:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NABULSI,T,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: NABULSI,T
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 16:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:46

SUBJECT'S LAST NAME: BURTON

DATE/TIME OF TEST: 05/12/2018 17:22:17

Diagnostics Check	OK	17:06	
Air Blank	0.000	17:06	
Control Test	0.080	17:07	
Air Blank	0.000	17:07	
Subject Sample	# 1 0.115	17:10	Volume Not Met
Breath Volume	0.046	17:10	
Air Blank	0.000	17:11	
Air Blank	0.000	17:13	
Subject Sample	# 2 0.152	17:16	
Breath Volume	1.132	17:16	
Air Blank	0.000	17:16	
Air Blank	0.000	17:18	
Subject Sample	# 3 0.158	17:20	
Breath Volume	1.718	17:20	
Air Blank	0.000	17:21	
Control Test	0.078	17:21	
Air Blank	0.000	17:22	
Diagnostics Check	OK	17:22	

Data Download Date/Time: 06/06/2018 08:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILHELM, TONY, J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: NABULSI, MOOTAZ
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 16:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: WISEMAN

DATE/TIME OF TEST: 05/13/2018 17:23:07

Diagnostics Check	OK	17:20	
Air Blank	0.000	17:20	
Control Test	0.080	17:21	
Air Blank	0.000	17:21	
Subject Sample	# 1	17:21	Subject Test Refused
Breath Volume	0.000	17:21	
Air Blank	0.000	17:22	
Control Test	0.081	17:22	
Air Blank	0.000	17:23	
Diagnostics Check	OK	17:23	

Data Download Date/Time: 06/06/2018 08:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,RAYMOND,S
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: MILLER,RAYMOND
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: LATHAM

DATE/TIME OF TEST: 05/17/2018 04:14:05

Diagnostics Check	OK	04:04
Air Blank	0.000	04:05
Control Test	0.080	04:05
Air Blank	0.000	04:05
Subject Sample # 1	0.150	04:09
Breath Volume	1.816	04:09
Air Blank	0.000	04:09
Air Blank	0.000	04:11
Subject Sample # 2	0.146	04:12
Breath Volume	1.632	04:12
Air Blank	0.000	04:13
Control Test	0.079	04:13
Air Blank	0.000	04:14
Diagnostics Check	OK	04:14

Data Download Date/Time: 06/06/2018 08:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANCHEZ, HECTOR, W
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SANCHEZ, HECTOR
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: GREGORY

DATE/TIME OF TEST: 05/17/2018 22:30:30

Diagnostics Check	OK	22:26	
Air Blank	0.000	22:26	
Control Test	0.080	22:26	
Air Blank	0.000	22:27	
Subject Sample	# 1	22:29	Subject Test Refused
Breath Volume	0.000	22:29	
Air Blank	0.000	22:29	
Control Test	0.080	22:29	
Air Blank	0.000	22:30	
Diagnostics Check	OK	22:30	

Data Download Date/Time: 06/06/2018 08:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FITCHETT, ADAM, R
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: HAMILTON, JOSHUA
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: VOJIK

DATE/TIME OF TEST: 05/18/2018 23:53:35

Diagnostics Check	OK	23:42	
Air Blank	0.000	23:42	
Control Test	0.080	23:42	
Air Blank	0.000	23:43	
Subject Sample	# 1	23:46	No Sample Provided
Breath Volume	0.000	23:46	
Air Blank	0.000	23:47	
Air Blank	0.000	23:49	
Subject Sample	# 2	23:52	No Sample Provided
Breath Volume	0.000	23:52	
Air Blank	0.000	23:52	
Control Test	0.079	23:52	
Air Blank	0.000	23:53	
Diagnostics Check	OK	23:53	

Data Download Date/Time: 06/06/2018 08:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORBITT,KYLE,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: CORBITT,KYLE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: BUI
ARREST TIME: 16:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:20

SUBJECT'S LAST NAME: MCCUTCHEON

DATE/TIME OF TEST: 05/19/2018 18:49:43

Diagnostics Check	OK	18:41
Air Blank	0.000	18:42
Control Test	0.080	18:42
Air Blank	0.000	18:43
Subject Sample # 1	0.207	18:45
Breath Volume	4.292	18:45
Air Blank	0.000	18:45
Air Blank	0.000	18:47
Subject Sample # 2	0.194	18:48
Breath Volume	1.816	18:48
Air Blank	0.000	18:48
Control Test	0.077	18:49
Air Blank	0.000	18:49
Diagnostics Check	OK	18:49

Data Download Date/Time: 06/06/2018 08:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIMBRIL,JOSHUA,B
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: KIMBRIL,JOSHUA
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 18:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: DEUELL

DATE/TIME OF TEST: 05/21/2018 19:58:45

Diagnostics Check	OK	19:50
Air Blank	0.000	19:51
Control Test	0.080	19:51
Air Blank	0.000	19:52
Subject Sample # 1	0.265	19:53
Breath Volume	2.824	19:53
Air Blank	0.000	19:54
Air Blank	0.000	19:55
Subject Sample # 2	0.269	19:57
Breath Volume	2.683	19:57
Air Blank	0.000	19:57
Control Test	0.079	19:58
Air Blank	0.000	19:58
Diagnostics Check	OK	19:58

Data Download Date/Time: 06/06/2018 08:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITMAN,AARON,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: COTTRELL,TORI
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 18:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: VAUGHAN

DATE/TIME OF TEST: 05/24/2018 19:17:21

Diagnostics Check	OK	19:14	
Air Blank	0.000	19:14	
Control Test	0.079	19:14	
Air Blank	0.000	19:15	
Subject Sample	# 1	19:15	Subject Test Refused
Breath Volume	0.000	19:15	
Air Blank	0.000	19:16	
Control Test	0.080	19:16	
Air Blank	0.000	19:17	
Diagnostics Check	OK	19:17	

Data Download Date/Time: 06/06/2018 08:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOGGAN,MARLON,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BOGGAN,MARLON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 17:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:31

SUBJECT'S LAST NAME: HOPPEL

DATE/TIME OF TEST: 05/25/2018 17:57:39

Diagnostics Check	OK	17:51
Air Blank	0.000	17:51
Control Test	0.080	17:52
Air Blank	0.000	17:52
Subject Sample # 1	0.000	17:53
Breath Volume	3.437	17:53
Air Blank	0.000	17:53
Air Blank	0.000	17:55
Subject Sample # 2	0.000	17:56
Breath Volume	4.660	17:56
Air Blank	0.000	17:56
Control Test	0.080	17:57
Air Blank	0.000	17:57
Diagnostics Check	OK	17:57

Data Download Date/Time: 06/06/2018 08:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIMBRIL,JOSHUA,B
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: KIMBRIL,JOSHUA
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 05/28/2018 03:57:34

Diagnostics Check	OK	03:50
Air Blank	0.000	03:51
Control Test	0.080	03:51
Air Blank	0.000	03:52
Subject Sample # 1	0.109	03:53
Breath Volume	4.230	03:53
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample # 2	0.106	03:56
Breath Volume	3.824	03:56
Air Blank	0.000	03:56
Control Test	0.080	03:56
Air Blank	0.000	03:57
Diagnostics Check	OK	03:57

Data Download Date/Time: 06/06/2018 08:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: THOMASON, TYLER, J

BREATH TEST OPERATOR AGENCY: Crestview PD

ARRESTING OFFICER: PEAK, JAY

ARRESTING OFFICER AGENCY: Crestview PD

LAST AGENCY INSPECTION: 04/26/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 31517080A1

DRY GAS EXPIRY DATE: 01/05/2020

VIOLATION CODE: DUI

ARREST TIME: 23:39

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: O'MALLEY

DATE/TIME OF TEST: 05/26/2018 00:12:14

Diagnostics Check	OK	00:05
Air Blank	0.000	00:05
Control Test	0.079	00:05
Air Blank	0.000	00:06
Subject Sample # 1	0.150	00:07
Breath Volume	1.363	00:07
Air Blank	0.000	00:07
Air Blank	0.000	00:09
Subject Sample # 2	0.139	00:10
Breath Volume	3.203	00:10
Air Blank	0.000	00:11
Control Test	0.076	00:11
Air Blank	0.000	00:12
Diagnostics Check	OK	00:12

Data Download Date/Time: 05/31/2018 17:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER, TYLER, J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KOSTAYL, M
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 04/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 05/11/2018 01:42:30

Diagnostics Check	OK	01:36
Air Blank	0.000	01:36
Control Test	0.077	01:36
Air Blank	0.000	01:37
Subject Sample # 1	0.176	01:38
Breath Volume	1.777	01:38
Air Blank	0.000	01:38
Air Blank	0.000	01:40
Subject Sample # 2	0.185	01:40
Breath Volume	2.023	01:40
Air Blank	0.000	01:41
Control Test	0.078	01:41
Air Blank	0.000	01:42
Diagnostics Check	OK	01:42

Data Download Date/Time: 05/31/2018 08:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005077

INSTRUMENT REGISTERED TO: NORTH MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PINEDA,MIGUEL,A
BREATH TEST OPERATOR AGENCY: North Miami PD
ARRESTING OFFICER: GONZALEZ,SCARLETT
ARRESTING OFFICER AGENCY: North Miami PD
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 09:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:43

SUBJECT'S LAST NAME: VERGARA

DATE/TIME OF TEST: 05/27/2018 10:12:52

Diagnostics Check	OK	10:06
Air Blank	0.000	10:06
Control Test	0.079	10:07
Air Blank	0.000	10:07
Subject Sample # 1	0.241	10:08
Breath Volume	2.765	10:08
Air Blank	0.000	10:09
Air Blank	0.000	10:10
Subject Sample # 2	0.236	10:11
Breath Volume	2.503	10:11
Air Blank	0.000	10:11
Control Test	0.079	10:12
Air Blank	0.000	10:12
Diagnostics Check	OK	10:12

Data Download Date/Time: 06/18/2018 13:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS, RONALD, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PAULINO, M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: CANTU-FIGUEROA

DATE/TIME OF TEST: 05/03/2018 22:53:17

Diagnostics Check	OK	22:46
Air Blank	0.000	22:47
Control Test	0.081	22:47
Air Blank	0.000	22:48
Subject Sample # 1	0.181	22:48
Breath Volume	2.222	22:48
Air Blank	0.000	22:49
Air Blank	0.000	22:51
Subject Sample # 2	0.177	22:51
Breath Volume	1.890	22:51
Air Blank	0.000	22:52
Control Test	0.081	22:52
Air Blank	0.000	22:53
Diagnostics Check	OK	22:53

Data Download Date/Time: 06/19/2018 19:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SINANI,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: LOPEZ-PEREZ

DATE/TIME OF TEST: 05/03/2018 23:34:53

Diagnostics Check	OK	23:27
Air Blank	0.000	23:28
Control Test	0.081	23:28
Air Blank	0.000	23:28
Subject Sample # 1	0.146	23:30
Breath Volume	2.617	23:30
Air Blank	0.000	23:30
Air Blank	0.000	23:32
Subject Sample # 2	0.150	23:33
Breath Volume	2.453	23:33
Air Blank	0.000	23:33
Control Test	0.081	23:34
Air Blank	0.000	23:34
Diagnostics Check	OK	23:34

Data Download Date/Time: 06/19/2018 19:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: CARREJO-GONZALEZ

DATE/TIME OF TEST: 05/04/2018 22:18:01

Diagnostics Check	OK	22:10
Air Blank	0.000	22:11
Control Test	0.081	22:11
Air Blank	0.000	22:12
Subject Sample # 1	0.195	22:13
Breath Volume	3.074	22:13
Air Blank	0.000	22:14
Air Blank	0.000	22:15
Subject Sample # 2	0.198	22:16
Breath Volume	2.609	22:16
Air Blank	0.000	22:17
Control Test	0.081	22:17
Air Blank	0.000	22:17
Diagnostics Check	OK	22:18

Data Download Date/Time: 06/19/2018 19:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS, RONALD, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NIEUWENDAM, R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 05/05/2018 00:52:56

Diagnostics Check	OK	00:46
Air Blank	0.000	00:47
Control Test	0.081	00:47
Air Blank	0.000	00:48
Subject Sample # 1	0.104	00:48
Breath Volume	3.113	00:48
Air Blank	0.000	00:49
Air Blank	0.000	00:50
Subject Sample # 2	0.103	00:51
Breath Volume	2.320	00:51
Air Blank	0.000	00:52
Control Test	0.080	00:52
Air Blank	0.000	00:52
Diagnostics Check	OK	00:52

Data Download Date/Time: 06/19/2018 19:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER, A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: EVERTS

DATE/TIME OF TEST: 05/05/2018 04:23:27

Diagnostics Check	OK	04:17
Air Blank	0.000	04:17
Control Test	0.081	04:18
Air Blank	0.000	04:18
Subject Sample # 1	0.173	04:19
Breath Volume	2.863	04:19
Air Blank	0.000	04:19
Air Blank	0.000	04:21
Subject Sample # 2	0.159	04:21
Breath Volume	1.250	04:21
Air Blank	0.000	04:22
Control Test	0.080	04:22
Air Blank	0.000	04:23
Diagnostics Check	OK	04:23

Data Download Date/Time: 06/19/2018 19:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIMPKINS,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:09

SUBJECT'S LAST NAME: VEGA VAZQUEZ

DATE/TIME OF TEST: 05/05/2018 05:36:54

Diagnostics Check	OK	05:29
Air Blank	0.000	05:29
Control Test	0.081	05:30
Air Blank	0.000	05:30
Subject Sample # 1	0.164	05:32
Breath Volume	1.785	05:32
Air Blank	0.000	05:32
Air Blank	0.000	05:34
Subject Sample # 2	0.166	05:35
Breath Volume	1.742	05:35
Air Blank	0.000	05:35
Control Test	0.080	05:36
Air Blank	0.000	05:36
Diagnostics Check	OK	05:36

Data Download Date/Time: 06/19/2018 19:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,REBECCA,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PECYLAK,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 06:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:59

SUBJECT'S LAST NAME: CANALES

DATE/TIME OF TEST: 05/05/2018 07:30:48

Diagnostics Check	OK	07:23
Air Blank	0.000	07:24
Control Test	0.080	07:24
Air Blank	0.000	07:25
Subject Sample # 1	0.145	07:26
Breath Volume	1.242	07:26
Air Blank	0.000	07:26
Air Blank	0.000	07:28
Subject Sample # 2	0.161	07:29
Breath Volume	3.003	07:29
Air Blank	0.000	07:29
Control Test	0.080	07:30
Air Blank	0.000	07:30
Diagnostics Check	OK	07:30

Data Download Date/Time: 06/19/2018 19:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,REBECCA,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: IVEY,C
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:11

SUBJECT'S LAST NAME: MINOR

DATE/TIME OF TEST: 05/05/2018 17:38:48

Diagnostics Check	OK	17:31
Air Blank	0.000	17:32
Control Test	0.081	17:32
Air Blank	0.000	17:33
Subject Sample # 1	0.189	17:34
Breath Volume	1.179	17:34
Air Blank	0.000	17:35
Air Blank	0.000	17:36
Subject Sample # 2	0.180	17:37
Breath Volume	1.140	17:37
Air Blank	0.000	17:37
Control Test	0.081	17:38
Air Blank	0.000	17:38
Diagnostics Check	OK	17:38

Data Download Date/Time: 06/19/2018 19:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: O'NEILL

DATE/TIME OF TEST: 05/06/2018 03:18:04

Diagnostics Check	OK	03:11
Air Blank	0.000	03:11
Control Test	0.082	03:12
Air Blank	0.000	03:12
Subject Sample # 1	0.157	03:13
Breath Volume	2.957	03:13
Air Blank	0.000	03:14
Air Blank	0.000	03:15
Subject Sample # 2	0.157	03:16
Breath Volume	2.660	03:16
Air Blank	0.000	03:17
Control Test	0.081	03:17
Air Blank	0.000	03:17
Diagnostics Check	OK	03:18

Data Download Date/Time: 06/19/2018 19:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: SOUTHARD

DATE/TIME OF TEST: 05/06/2018 03:43:45

Diagnostics Check	OK	03:36
Air Blank	0.000	03:36
Control Test	0.082	03:37
Air Blank	0.000	03:37
Subject Sample # 1	0.125	03:38
Breath Volume	1.261	03:38
Air Blank	0.000	03:39
Air Blank	0.000	03:41
Subject Sample # 2	0.125	03:42
Breath Volume	1.132	03:42
Air Blank	0.000	03:42
Control Test	0.081	03:43
Air Blank	0.000	03:43
Diagnostics Check	OK	03:43

Data Download Date/Time: 06/19/2018 19:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GREEN,W
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: MOLLINEDO VELIZ

DATE/TIME OF TEST: 05/06/2018 04:05:15

Diagnostics Check	OK	03:59
Air Blank	0.000	03:59
Control Test	0.081	03:59
Air Blank	0.000	04:00
Subject Sample # 1	0.079	04:00
Breath Volume	2.585	04:00
Air Blank	0.000	04:01
Air Blank	0.000	04:03
Subject Sample # 2	0.077	04:03
Breath Volume	3.117	04:03
Air Blank	0.000	04:04
Control Test	0.081	04:04
Air Blank	0.000	04:05
Diagnostics Check	OK	04:05

Data Download Date/Time: 06/19/2018 19:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NIEUWENDAM,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: KIBBY

DATE/TIME OF TEST: 05/06/2018 04:40:52

Diagnostics Check	OK	04:31
Air Blank	0.000	04:31
Control Test	0.081	04:32
Air Blank	0.000	04:32
Subject Sample # 1	0.179	04:35
Breath Volume	1.933	04:35
Air Blank	0.000	04:35
Air Blank	0.000	04:37
Subject Sample # 2	0.170	04:39
Breath Volume	1.457	04:39
Air Blank	0.000	04:39
Control Test	0.082	04:40
Air Blank	0.000	04:40
Diagnostics Check	OK	04:40

Data Download Date/Time: 06/19/2018 19:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: COSTA,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 05/06/2018 05:22:58

Diagnostics Check	OK	05:10	
Air Blank	0.000	05:11	
Control Test	0.081	05:11	
Air Blank	0.000	05:12	
Subject Sample	# 1 0.151	05:15	Volume Not Met
Breath Volume	0.097	05:15	
Air Blank	0.000	05:15	
Air Blank	0.000	05:17	
Subject Sample	# 2 0.154	05:18	
Breath Volume	1.265	05:18	
Air Blank	0.000	05:18	
Air Blank	0.000	05:20	
Subject Sample	# 3 0.153	05:21	
Breath Volume	1.406	05:21	
Air Blank	0.000	05:22	
Control Test	0.081	05:22	
Air Blank	0.000	05:22	
Diagnostics Check	OK	05:22	

Data Download Date/Time: 06/19/2018 19:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIMPKINS,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:48

SUBJECT'S LAST NAME: LEON AGUILAR

DATE/TIME OF TEST: 05/06/2018 21:16:42

Diagnostics Check	OK	21:08
Air Blank	0.000	21:08
Control Test	0.080	21:08
Air Blank	0.000	21:09
Subject Sample # 1	0.160	21:12
Breath Volume	1.242	21:12
Air Blank	0.000	21:12
Air Blank	0.000	21:14
Subject Sample # 2	0.154	21:15
Breath Volume	1.179	21:15
Air Blank	0.000	21:15
Control Test	0.081	21:16
Air Blank	0.000	21:16
Diagnostics Check	OK	21:16

Data Download Date/Time: 06/19/2018 19:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: ABSALOM

DATE/TIME OF TEST: 05/07/2018 02:54:06

Diagnostics Check	OK	02:50	
Air Blank	0.000	02:51	
Control Test	0.081	02:51	
Air Blank	0.000	02:52	
Subject Sample	# 1	02:52	Subject Test Refused
Breath Volume	0.000	02:52	
Air Blank	0.000	02:53	
Control Test	0.081	02:53	
Air Blank	0.000	02:54	
Diagnostics Check	OK	02:54	

Data Download Date/Time: 06/19/2018 19:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MOREJON,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:23

SUBJECT'S LAST NAME: MARTINEZ-GARCIA

DATE/TIME OF TEST: 05/07/2018 05:49:48

Diagnostics Check	OK	05:43
Air Blank	0.000	05:44
Control Test	0.080	05:44
Air Blank	0.000	05:44
Subject Sample # 1	0.273	05:45
Breath Volume	2.781	05:45
Air Blank	0.000	05:46
Air Blank	0.000	05:47
Subject Sample # 2	0.267	05:48
Breath Volume	2.539	05:48
Air Blank	0.000	05:48
Control Test	0.080	05:49
Air Blank	0.000	05:49
Diagnostics Check	OK	05:49

Data Download Date/Time: 06/19/2018 19:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,KAMIKA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SCHILLER,E
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: GULNICK

DATE/TIME OF TEST: 05/08/2018 23:22:20

Diagnostics Check	OK	23:15
Air Blank	0.000	23:16
Control Test	0.081	23:16
Air Blank	0.000	23:17
Subject Sample # 1	0.195	23:17
Breath Volume	3.464	23:17
Air Blank	0.000	23:18
Air Blank	0.000	23:20
Subject Sample # 2	0.195	23:20
Breath Volume	2.625	23:20
Air Blank	0.000	23:21
Control Test	0.080	23:21
Air Blank	0.000	23:22
Diagnostics Check	OK	23:22

Data Download Date/Time: 06/19/2018 19:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,KAMIKA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MACIAS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: SONZOGNI

DATE/TIME OF TEST: 05/09/2018 03:28:09

Diagnostics Check	OK	03:22
Air Blank	0.000	03:22
Control Test	0.081	03:23
Air Blank	0.000	03:23
Subject Sample # 1	0.000	03:23
Breath Volume	1.964	03:23
Air Blank	0.000	03:24
Air Blank	0.000	03:26
Subject Sample # 2	0.000	03:26
Breath Volume	2.335	03:26
Air Blank	0.000	03:27
Control Test	0.081	03:27
Air Blank	0.000	03:28
Diagnostics Check	OK	03:28

Data Download Date/Time: 06/19/2018 19:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS, RONALD, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DIEL, A
ARRESTING OFFICER AGENCY: U of So FL Campus PD
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: LUGO-NAVARRO

DATE/TIME OF TEST: 05/09/2018 23:58:39

Diagnostics Check	OK	23:52
Air Blank	0.000	23:53
Control Test	0.081	23:53
Air Blank	0.000	23:54
Subject Sample # 1	0.008	23:54
Breath Volume	4.175	23:54
Air Blank	0.000	23:54
Air Blank	0.000	23:56
Subject Sample # 2	0.007	23:57
Breath Volume	4.253	23:57
Air Blank	0.000	23:57
Control Test	0.081	23:58
Air Blank	0.000	23:58
Diagnostics Check	OK	23:58

Data Download Date/Time: 06/19/2018 19:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,KAMIKA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DUNCAN,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: BAGINSKI

DATE/TIME OF TEST: 05/10/2018 02:23:57

Diagnostics Check	OK	02:17
Air Blank	0.000	02:18
Control Test	0.081	02:18
Air Blank	0.000	02:18
Subject Sample # 1	0.194	02:19
Breath Volume	3.562	02:19
Air Blank	0.000	02:20
Air Blank	0.000	02:21
Subject Sample # 2	0.192	02:22
Breath Volume	3.757	02:22
Air Blank	0.000	02:23
Control Test	0.081	02:23
Air Blank	0.000	02:23
Diagnostics Check	OK	02:23

Data Download Date/Time: 06/19/2018 19:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: WILLIAMS II

DATE/TIME OF TEST: 05/11/2018 00:58:27

Diagnostics Check	OK	00:44	
Air Blank	0.000	00:44	
Control Test	0.081	00:44	
Air Blank	0.000	00:45	
Subject Sample	# 1 0.252	00:48	Volume Not Met
Breath Volume	0.757	00:48	
Air Blank	0.000	00:49	
Air Blank	0.000	00:50	
Subject Sample	# 2 0.254	00:52	
Breath Volume	1.207	00:52	
Air Blank	0.000	00:53	
Air Blank	0.000	00:54	
Subject Sample	# 3 0.251	00:56	
Breath Volume	1.347	00:56	
Air Blank	0.000	00:57	
Control Test	0.080	00:57	
Air Blank	0.000	00:58	
Diagnostics Check	OK	00:58	

Data Download Date/Time: 06/19/2018 19:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HILES,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: RAMOS-RODRIGUEZ

DATE/TIME OF TEST: 05/11/2018 04:33:38

Diagnostics Check	OK	04:27
Air Blank	0.000	04:27
Control Test	0.081	04:27
Air Blank	0.000	04:28
Subject Sample # 1	0.149	04:28
Breath Volume	1.929	04:28
Air Blank	0.000	04:29
Air Blank	0.000	04:31
Subject Sample # 2	0.145	04:32
Breath Volume	1.214	04:32
Air Blank	0.000	04:32
Control Test	0.081	04:33
Air Blank	0.000	04:33
Diagnostics Check	OK	04:33

Data Download Date/Time: 06/19/2018 19:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SCHILLER,E
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: MASSEY

DATE/TIME OF TEST: 05/11/2018 23:41:41

Diagnostics Check	OK	23:34
Air Blank	0.000	23:35
Control Test	0.082	23:35
Air Blank	0.000	23:36
Subject Sample # 1	0.185	23:37
Breath Volume	2.441	23:37
Air Blank	0.000	23:37
Air Blank	0.000	23:39
Subject Sample # 2	0.195	23:40
Breath Volume	1.593	23:40
Air Blank	0.000	23:40
Control Test	0.080	23:41
Air Blank	0.000	23:41
Diagnostics Check	OK	23:41

Data Download Date/Time: 06/19/2018 19:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRANKLIN,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: CRUICKSHANK

DATE/TIME OF TEST: 05/12/2018 01:14:26

Diagnostics Check	OK	01:08
Air Blank	0.000	01:08
Control Test	0.081	01:08
Air Blank	0.000	01:09
Subject Sample # 1	0.128	01:09
Breath Volume	1.367	01:09
Air Blank	0.000	01:10
Air Blank	0.000	01:12
Subject Sample # 2	0.146	01:12
Breath Volume	1.804	01:12
Air Blank	0.000	01:13
Control Test	0.082	01:13
Air Blank	0.000	01:14
Diagnostics Check	OK	01:14

Data Download Date/Time: 06/19/2018 19:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 05/12/2018 03:47:24

Diagnostics Check	OK	03:31	
Air Blank	0.000	03:32	
Control Test	0.081	03:32	
Air Blank	0.000	03:32	
Subject Sample	# 1	03:36	No Sample Provided
Breath Volume	0.000	03:36	
Air Blank	0.000	03:36	
Air Blank	0.000	03:38	
Subject Sample	# 2	0.166	03:40
Breath Volume	1.128	03:40	
Air Blank	0.000	03:41	
Air Blank	0.000	03:42	
Subject Sample	# 3	0.171	03:45
Breath Volume	1.589	03:45	
Air Blank	0.000	03:46	
Control Test	0.083	03:46	
Air Blank	0.000	03:47	
Diagnostics Check	OK	03:47	

Data Download Date/Time: 06/19/2018 19:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: BEALL

DATE/TIME OF TEST: 05/12/2018 05:01:21

Diagnostics Check	OK	04:53
Air Blank	0.000	04:53
Control Test	0.081	04:53
Air Blank	0.000	04:54
Subject Sample # 1	0.087	04:56
Breath Volume	1.242	04:56
Air Blank	0.000	04:56
Air Blank	0.000	04:58
Subject Sample # 2	0.090	04:59
Breath Volume	2.660	04:59
Air Blank	0.000	05:00
Control Test	0.082	05:00
Air Blank	0.000	05:01
Diagnostics Check	OK	05:01

Data Download Date/Time: 06/19/2018 19:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN, T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: WARD

DATE/TIME OF TEST: 05/12/2018 19:13:57

Diagnostics Check	OK	19:11	
Air Blank	0.000	19:11	
Control Test	0.081	19:11	
Air Blank	0.000	19:12	
Subject Sample	# 1	19:12	Subject Test Refused
Breath Volume	0.000	19:12	
Air Blank	0.000	19:13	
Control Test	0.081	19:13	
Air Blank	0.000	19:13	
Diagnostics Check	OK	19:13	

Data Download Date/Time: 06/19/2018 19:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,KAMIKA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RORER,J
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: BELISLE

DATE/TIME OF TEST: 05/12/2018 22:21:07

Diagnostics Check	OK	22:17	
Air Blank	0.000	22:17	
Control Test	0.082	22:17	
Air Blank	0.000	22:18	
Subject Sample	# 1	22:19	Subject Test Refused
Breath Volume	0.000	22:19	
Air Blank	0.000	22:20	
Control Test	0.081	22:20	
Air Blank	0.000	22:21	
Diagnostics Check	OK	22:21	

Data Download Date/Time: 06/19/2018 19:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ROWLES,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: ADMIN
ARREST TIME: 18:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: SANDERS

DATE/TIME OF TEST: 05/14/2018 19:32:29

Diagnostics Check	OK	19:27	
Air Blank	0.000	19:28	
Control Test	0.082	19:28	
Air Blank	0.000	19:28	
Subject Sample	# 1 0.049	19:29	
Breath Volume	3.664	19:29	
Air Blank	0.000	19:29	
Air Blank	0.000	19:31	
Subject Sample	# 2	19:32	Sequence Aborted
Breath Volume	0.000	19:32	
Air Blank	0.000	19:32	

Data Download Date/Time: 06/19/2018 19:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKIPPER,ROGER,G
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VELAZQUEZ,B 446
ARRESTING OFFICER AGENCY: U of So FL Campus PD
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 09:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:03

SUBJECT'S LAST NAME: KRUSA

DATE/TIME OF TEST: 05/15/2018 11:31:27

Diagnostics Check	OK	11:24
Air Blank	0.000	11:25
Control Test	0.081	11:25
Air Blank	0.000	11:26
Subject Sample # 1	0.019	11:26
Breath Volume	2.582	11:26
Air Blank	0.000	11:27
Air Blank	0.000	11:29
Subject Sample # 2	0.016	11:29
Breath Volume	1.769	11:29
Air Blank	0.000	11:30
Control Test	0.081	11:30
Air Blank	0.000	11:31
Diagnostics Check	OK	11:31

Data Download Date/Time: 06/19/2018 19:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005248

INSTRUMENT REGISTERED TO: FISH & WILDLIFE CC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLANCO,JUAN,S
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: BLANCO,JUAN
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 04/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: BUI
ARREST TIME: 21:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 05/06/2018 21:58:32

Diagnostics Check	OK	21:53	
Air Blank	0.000	21:53	
Control Test	0.080	21:54	
Air Blank	0.000	21:54	
Subject Sample	# 1	21:57	Subject Test Refused
Breath Volume	0.000	21:57	
Air Blank	0.000	21:57	
Control Test	0.080	21:57	
Air Blank	0.000	21:58	
Diagnostics Check	OK	21:58	

Data Download Date/Time: 05/31/2018 11:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO,I,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: VALLE,D
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A3
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: FARBER

DATE/TIME OF TEST: 05/04/2018 22:14:30

Diagnostics Check	OK	22:08
Air Blank	0.000	22:08
Control Test	0.080	22:08
Air Blank	0.000	22:09
Subject Sample # 1	0.265	22:09
Breath Volume	2.128	22:09
Air Blank	0.000	22:10
Air Blank	0.000	22:12
Subject Sample # 2	0.254	22:12
Breath Volume	1.281	22:12
Air Blank	0.000	22:13
Control Test	0.079	22:13
Air Blank	0.000	22:14
Diagnostics Check	OK	22:14

Data Download Date/Time: 05/13/2018 08:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTA,ANGEL,L
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: ORTIZ,B
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A3
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: LINTON

DATE/TIME OF TEST: 05/08/2018 22:01:57

Diagnostics Check	OK	21:58	
Air Blank	0.000	21:59	
Control Test	0.079	21:59	
Air Blank	0.000	22:00	
Subject Sample	# 1	22:00	Subject Test Refused
Breath Volume	0.000	22:00	
Air Blank	0.000	22:01	
Control Test	0.079	22:01	
Air Blank	0.000	22:01	
Diagnostics Check	OK	22:01	

Data Download Date/Time: 05/13/2018 08:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARTNER,ARTHUR,F
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GARTNER,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A3
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: BARTOLIS

DATE/TIME OF TEST: 05/10/2018 03:32:29

Diagnostics Check	OK	03:26
Air Blank	0.000	03:26
Control Test	0.079	03:26
Air Blank	0.000	03:27
Subject Sample # 1	0.131	03:28
Breath Volume	2.812	03:28
Air Blank	0.000	03:28
Air Blank	0.000	03:30
Subject Sample # 2	0.130	03:30
Breath Volume	2.281	03:30
Air Blank	0.000	03:31
Control Test	0.078	03:31
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

Data Download Date/Time: 05/13/2018 08:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARTNER,ARTHUR,F
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GARTNER,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A3
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 12:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: GENDRON

DATE/TIME OF TEST: 05/12/2018 01:32:36

Diagnostics Check	OK	01:25
Air Blank	0.000	01:26
Control Test	0.079	01:26
Air Blank	0.000	01:26
Subject Sample # 1	0.110	01:27
Breath Volume	1.910	01:27
Air Blank	0.000	01:28
Air Blank	0.000	01:30
Subject Sample # 2	0.113	01:31
Breath Volume	1.941	01:31
Air Blank	0.000	01:31
Control Test	0.078	01:32
Air Blank	0.000	01:32
Diagnostics Check	OK	01:32

Data Download Date/Time: 05/13/2018 08:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASCITELLI,V,D
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MASCITELLI,V
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 05/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A3
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: BLODGETT

DATE/TIME OF TEST: 05/20/2018 03:34:16

Diagnostics Check	OK	03:20	
Air Blank	0.000	03:21	
Control Test	0.078	03:21	
Air Blank	0.000	03:22	
Subject Sample	# 1	03:25	No Sample Provided
Breath Volume	0.000	03:25	
Air Blank	0.000	03:25	
Air Blank	0.000	03:27	
Subject Sample	# 2	0.145	03:29
Breath Volume	1.417	03:29	
Air Blank	0.000	03:30	
Air Blank	0.000	03:32	
Subject Sample	# 3	0.150	03:32
Breath Volume	1.667	03:32	
Air Blank	0.000	03:33	
Control Test	0.078	03:33	
Air Blank	0.000	03:34	
Diagnostics Check	OK	03:34	

Data Download Date/Time: 06/23/2018 08:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO,I,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: SHUTT,T
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 05/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A3
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:51

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 05/25/2018 00:18:22

Diagnostics Check	OK	00:11
Air Blank	0.000	00:11
Control Test	0.080	00:12
Air Blank	0.000	00:12
Subject Sample # 1	0.177	00:13
Breath Volume	2.828	00:13
Air Blank	0.000	00:14
Air Blank	0.000	00:15
Subject Sample # 2	0.176	00:16
Breath Volume	3.781	00:16
Air Blank	0.000	00:17
Control Test	0.079	00:17
Air Blank	0.000	00:18
Diagnostics Check	OK	00:18

Data Download Date/Time: 06/23/2018 08:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STIMPSON,RACHEL,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: GREIS-OSPINA,DAVID
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 05/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A3
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 13:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:30

SUBJECT'S LAST NAME: GLEW

DATE/TIME OF TEST: 05/31/2018 13:53:53

Diagnostics Check	OK	13:50	
Air Blank	0.000	13:51	
Control Test	0.079	13:51	
Air Blank	0.000	13:52	
Subject Sample	# 1	13:52	Subject Test Refused
Breath Volume	0.000	13:52	
Air Blank	0.000	13:52	
Control Test	0.079	13:53	
Air Blank	0.000	13:53	
Diagnostics Check	OK	13:53	

Data Download Date/Time: 06/23/2018 08:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BREGON,JAMES,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BREGON,JAMES
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 04:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:04

SUBJECT'S LAST NAME: SEBASTIANO

DATE/TIME OF TEST: 05/01/2018 05:26:42

Diagnostics Check	OK	05:25	
Air Blank	0.000	05:25	
Control Test	0.081	05:25	
Air Blank		05:26	RFI Detect
Air Blank	0.000	05:26	

Data Download Date/Time: 05/30/2018 21:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BREGON,JAMES,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BREGON,JAMES
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 04:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:04

SUBJECT'S LAST NAME: SEBASTIANO

DATE/TIME OF TEST: 05/01/2018 05:36:15

Diagnostics Check	OK	05:29
Air Blank	0.000	05:30
Control Test	0.080	05:30
Air Blank	0.000	05:30
Subject Sample # 1	0.143	05:31
Breath Volume	2.527	05:31
Air Blank	0.000	05:32
Air Blank	0.000	05:34
Subject Sample # 2	0.162	05:34
Breath Volume	2.007	05:34
Air Blank	0.000	05:35
Control Test	0.080	05:35
Air Blank	0.000	05:36
Diagnostics Check	OK	05:36

Data Download Date/Time: 05/30/2018 21:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PANIAGUA,LT MICHAEL,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PANIAGUA,LT MICHAEL
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:16

SUBJECT'S LAST NAME: BREED

DATE/TIME OF TEST: 05/01/2018 22:43:15

Diagnostics Check	OK	22:36
Air Blank	0.000	22:37
Control Test	0.081	22:37
Air Blank	0.000	22:38
Subject Sample # 1	0.263	22:38
Breath Volume	2.867	22:38
Air Blank	0.000	22:39
Air Blank	0.000	22:40
Subject Sample # 2	0.262	22:41
Breath Volume	1.550	22:41
Air Blank	0.000	22:42
Control Test	0.081	22:42
Air Blank	0.000	22:43
Diagnostics Check	OK	22:43

Data Download Date/Time: 05/30/2018 21:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORR,RANDALL,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DORR,RANDALL
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: ALYEA

DATE/TIME OF TEST: 05/01/2018 23:49:29

Diagnostics Check	OK	23:42
Air Blank	0.000	23:42
Control Test	0.080	23:43
Air Blank	0.000	23:43
Subject Sample # 1	0.122	23:44
Breath Volume	1.121	23:44
Air Blank	0.000	23:45
Air Blank	0.000	23:47
Subject Sample # 2	0.121	23:47
Breath Volume	2.269	23:47
Air Blank	0.000	23:48
Control Test	0.080	23:48
Air Blank	0.000	23:49
Diagnostics Check	OK	23:49

Data Download Date/Time: 05/30/2018 21:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MARTINEZ,ALEX
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 04/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 05/02/2018 19:36:35

Diagnostics Check	OK	19:30	
Air Blank	0.000	19:31	
Control Test	0.081	19:31	
Air Blank	0.000	19:31	
Subject Sample	# 1 0.084	19:32	
Breath Volume	1.492	19:32	
Air Blank	0.000	19:32	
Air Blank	0.000	19:34	
Subject Sample	# 2 0.089	19:35	
Breath Volume	1.664	19:35	
Air Blank	0.000	19:35	
Control Test	0.080	19:36	
Air Blank		19:36	RFI Detect
Air Blank	0.000	19:36	

Data Download Date/Time: 05/30/2018 21:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MARTINEZ,ALEX
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 04/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 05/02/2018 19:45:02

Diagnostics Check	OK	19:39
Air Blank	0.000	19:39
Control Test	0.080	19:39
Air Blank	0.000	19:40
Subject Sample # 1	0.081	19:40
Breath Volume	1.593	19:40
Air Blank	0.000	19:41
Air Blank	0.000	19:43
Subject Sample # 2	0.083	19:43
Breath Volume	1.597	19:43
Air Blank	0.000	19:44
Control Test	0.080	19:44
Air Blank	0.000	19:44
Diagnostics Check	OK	19:45

Data Download Date/Time: 05/30/2018 21:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: KISS

DATE/TIME OF TEST: 05/02/2018 23:01:04

Diagnostics Check	OK	22:58	
Air Blank	0.000	22:58	
Control Test	0.081	22:59	
Air Blank	0.000	22:59	
Subject Sample	# 1	22:59	Subject Test Refused
Breath Volume	0.000	22:59	
Air Blank	0.000	23:00	
Control Test	0.081	23:00	
Air Blank	0.000	23:00	
Diagnostics Check	OK	23:01	

Data Download Date/Time: 05/30/2018 21:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FEIDT,E
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: THANDALAM MADAPUI SR

DATE/TIME OF TEST: 05/12/2018 00:52:08

Diagnostics Check	OK	00:44
Air Blank	0.000	00:45
Control Test	0.081	00:45
Air Blank	0.000	00:45
Subject Sample # 1	0.166	00:46
Breath Volume	1.316	00:46
Air Blank	0.000	00:47
Air Blank	0.000	00:48
Subject Sample # 2	0.176	00:50
Breath Volume	1.539	00:50
Air Blank	0.000	00:51
Control Test	0.080	00:51
Air Blank	0.000	00:52
Diagnostics Check	OK	00:52

Data Download Date/Time: 05/30/2018 21:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMIN,N
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: HUNDLEY

DATE/TIME OF TEST: 05/12/2018 02:32:24

Diagnostics Check	OK	02:29	
Air Blank	0.000	02:30	
Control Test	0.080	02:30	
Air Blank	0.000	02:30	
Subject Sample	# 1	02:31	Subject Test Refused
Breath Volume	0.000	02:31	
Air Blank	0.000	02:31	
Control Test	0.080	02:31	
Air Blank	0.000	02:32	
Diagnostics Check	OK	02:32	

Data Download Date/Time: 05/30/2018 21:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTHONY, MICHAEL,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT, NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: GOLSON

DATE/TIME OF TEST: 05/28/2018 23:25:43

Diagnostics Check	OK	23:22	
Air Blank	0.000	23:23	
Control Test	0.080	23:23	
Air Blank	0.000	23:24	
Subject Sample	# 1	23:24	Subject Test Refused
Breath Volume	0.000	23:24	
Air Blank	0.000	23:24	
Control Test	0.079	23:25	
Air Blank	0.000	23:25	
Diagnostics Check	OK	23:25	

Data Download Date/Time: 05/30/2018 21:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWINDLE,CHARLES,M
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: SWINDLE,CHARLES
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 04/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: DUI
ARREST TIME: 17:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: ALEXANDER

DATE/TIME OF TEST: 05/14/2018 18:11:14

Diagnostics Check	OK	18:06	
Air Blank	0.000	18:06	
Control Test	0.081	18:06	
Air Blank	0.000	18:07	
Subject Sample	# 1	18:09	Subject Test Refused
Breath Volume	0.195	18:09	
Air Blank	0.000	18:10	
Control Test	0.083	18:10	
Air Blank	0.000	18:11	
Diagnostics Check	OK	18:11	

Data Download Date/Time: 05/25/2018 17:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICKETSON,JAY,R
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: GUY,M
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 05/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: DUI
ARREST TIME: 11:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: WILES

DATE/TIME OF TEST: 05/27/2018 00:22:57

Diagnostics Check	OK	00:20	
Air Blank	0.000	00:20	
Control Test	0.081	00:20	
Air Blank	0.000	00:21	
Subject Sample	# 1	00:21	Subject Test Refused
Breath Volume	0.000	00:21	
Air Blank	0.000	00:22	
Control Test	0.081	00:22	
Air Blank	0.000	00:22	
Diagnostics Check	OK	00:22	

Data Download Date/Time: 06/29/2018 11:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOULD, LEONARD,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: DAVIS, LEONARD
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: SCHENK

DATE/TIME OF TEST: 05/13/2018 23:47:19

Diagnostics Check	OK	23:40
Air Blank	0.000	23:41
Control Test	0.081	23:41
Air Blank	0.000	23:42
Subject Sample # 1	0.148	23:42
Breath Volume	2.109	23:42
Air Blank	0.000	23:43
Air Blank	0.000	23:45
Subject Sample # 2	0.149	23:45
Breath Volume	2.582	23:45
Air Blank	0.000	23:46
Control Test	0.080	23:46
Air Blank	0.000	23:47
Diagnostics Check	OK	23:47

Data Download Date/Time: 06/06/2018 10:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,EARTHEN,A
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: BROWN,EARTHEN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: STONEQUIST

DATE/TIME OF TEST: 05/14/2018 02:50:23

Diagnostics Check	OK	02:44
Air Blank	0.000	02:44
Control Test	0.081	02:44
Air Blank	0.000	02:45
Subject Sample # 1	0.199	02:45
Breath Volume	2.363	02:45
Air Blank	0.000	02:46
Air Blank	0.000	02:48
Subject Sample # 2	0.196	02:48
Breath Volume	2.140	02:48
Air Blank	0.000	02:49
Control Test	0.080	02:49
Air Blank	0.000	02:50
Diagnostics Check	OK	02:50

Data Download Date/Time: 06/06/2018 10:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,CHRISTINE,K
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: HALL,CHRISTINE
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 05/15/2018 23:26:08

Diagnostics Check	OK	23:19
Air Blank	0.000	23:19
Control Test	0.081	23:19
Air Blank	0.000	23:20
Subject Sample # 1	0.179	23:21
Breath Volume	2.300	23:21
Air Blank	0.000	23:22
Air Blank	0.000	23:23
Subject Sample # 2	0.170	23:24
Breath Volume	2.109	23:24
Air Blank	0.000	23:25
Control Test	0.079	23:25
Air Blank	0.000	23:26
Diagnostics Check	OK	23:26

Data Download Date/Time: 06/06/2018 10:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,JAMES,T
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: JOSEPH,JEAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:47

SUBJECT'S LAST NAME: SANDERS

DATE/TIME OF TEST: 05/17/2018 20:21:34

Diagnostics Check	OK	20:14
Air Blank	0.000	20:15
Control Test	0.080	20:15
Air Blank	0.000	20:15
Subject Sample # 1	0.199	20:16
Breath Volume	1.304	20:16
Air Blank	0.000	20:17
Air Blank	0.000	20:18
Subject Sample # 2	0.191	20:19
Breath Volume	1.246	20:19
Air Blank	0.000	20:20
Control Test	0.077	20:20
Air Blank	0.000	20:21
Diagnostics Check	OK	20:21

Data Download Date/Time: 06/06/2018 10:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,EARTHEN,A
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: BROWN,EARTHEN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: ZERBE

DATE/TIME OF TEST: 05/18/2018 03:23:27

Diagnostics Check	OK	03:16
Air Blank	0.000	03:17
Control Test	0.080	03:17
Air Blank	0.000	03:18
Subject Sample # 1	0.211	03:18
Breath Volume	2.730	03:18
Air Blank	0.000	03:19
Air Blank	0.000	03:21
Subject Sample # 2	0.198	03:21
Breath Volume	2.027	03:21
Air Blank	0.000	03:22
Control Test	0.080	03:22
Air Blank	0.000	03:23
Diagnostics Check	OK	03:23

Data Download Date/Time: 06/06/2018 10:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,EARTHEN,A
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: PAHLKE,M
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: DAWKINS

DATE/TIME OF TEST: 05/18/2018 04:01:41

Diagnostics Check	OK	03:55
Air Blank	0.000	03:55
Control Test	0.080	03:55
Air Blank	0.000	03:56
Subject Sample # 1	0.202	03:56
Breath Volume	2.421	03:56
Air Blank	0.000	03:57
Air Blank	0.000	03:59
Subject Sample # 2	0.196	03:59
Breath Volume	1.992	03:59
Air Blank	0.000	04:00
Control Test	0.079	04:00
Air Blank	0.000	04:01
Diagnostics Check	OK	04:01

Data Download Date/Time: 06/06/2018 10:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,CHRISTINE,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: RIVERA,JEVIKA
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 05/19/2018 23:09:57

Diagnostics Check	OK	23:00	
Air Blank	0.000	23:00	
Control Test	0.081	23:01	
Air Blank	0.000	23:01	
Subject Sample	# 1 0.151	23:03	
Breath Volume	1.218	23:03	
Air Blank	0.000	23:04	
Air Blank	0.000	23:05	
Subject Sample	# 2 0.128	23:08	Volume Not Met
Breath Volume	0.761	23:08	
Air Blank		23:09	Purge Fail
Air Blank	0.000	23:09	

Data Download Date/Time: 06/06/2018 10:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,CHRISTINE,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: RIVERA,JEVIKA
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 05/19/2018 23:13:17

Diagnostics Check	OK	23:11	
Air Blank	0.000	23:12	
Control Test	0.072	23:12	Control Outside Tolerance
Air Blank	0.000	23:13	

Data Download Date/Time: 06/06/2018 10:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARGO,TIMOTHY,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: WARGO,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: SHELLY

DATE/TIME OF TEST: 05/20/2018 03:39:17

Diagnostics Check	OK	03:31
Air Blank	0.000	03:32
Control Test	0.080	03:32
Air Blank	0.000	03:32
Subject Sample # 1	0.181	03:34
Breath Volume	2.421	03:34
Air Blank	0.000	03:35
Air Blank	0.000	03:36
Subject Sample # 2	0.174	03:37
Breath Volume	1.343	03:37
Air Blank	0.000	03:38
Control Test	0.079	03:38
Air Blank	0.000	03:39
Diagnostics Check	OK	03:39

Data Download Date/Time: 06/06/2018 10:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,EARTHEN,A
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: BROWN,EARTHEN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 05/22/2018 23:36:34

Diagnostics Check	OK	23:33	
Air Blank	0.000	23:33	
Control Test	0.080	23:33	
Air Blank	0.000	23:34	
Subject Sample	# 1 0.383	23:35	
Breath Volume	2.542	23:35	
Air Blank		23:36	Purge Fail
Air Blank	0.000	23:36	

Data Download Date/Time: 06/06/2018 10:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROMAN,A,C
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: DAVIS,D
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: GARDNER

DATE/TIME OF TEST: 05/23/2018 05:42:44

Diagnostics Check	OK	05:35
Air Blank	0.000	05:35
Control Test	0.080	05:36
Air Blank	0.000	05:36
Subject Sample # 1	0.164	05:37
Breath Volume	3.585	05:37
Air Blank	0.000	05:38
Air Blank	0.000	05:39
Subject Sample # 2	0.167	05:40
Breath Volume	2.121	05:40
Air Blank	0.000	05:41
Control Test	0.079	05:42
Air Blank	0.000	05:42
Diagnostics Check	OK	05:42

Data Download Date/Time: 06/06/2018 10:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,J,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: RIVERA,J
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:33

SUBJECT'S LAST NAME: SALAZAR-LOPEZ

DATE/TIME OF TEST: 05/26/2018 05:13:58

Diagnostics Check	OK	04:58	
Air Blank	0.000	04:59	
Control Test	0.081	04:59	
Air Blank	0.000	05:00	
Subject Sample	# 1 0.155	05:03	Volume Not Met
Breath Volume	0.468	05:03	
Air Blank	0.000	05:04	
Air Blank	0.000	05:05	
Subject Sample	# 2 0.149	05:06	
Breath Volume	1.398	05:06	
Air Blank	0.000	05:07	
Air Blank	0.000	05:09	
Subject Sample	# 3	05:11	Subject Test Refused
Breath Volume	0.648	05:11	
Air Blank	0.000	05:12	
Control Test	0.078	05:13	
Air Blank	0.000	05:13	
Diagnostics Check	OK	05:13	

Data Download Date/Time: 06/06/2018 10:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,J,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: RIVERA,J
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:33

SUBJECT'S LAST NAME: SALAZAR-LOPEZ

DATE/TIME OF TEST: 05/26/2018 05:22:52

Diagnostics Check	OK	05:15
Air Blank	0.000	05:16
Control Test	0.075	05:16
Air Blank	0.000	05:16
Subject Sample # 1	0.149	05:18
Breath Volume	1.144	05:18
Air Blank	0.000	05:18
Air Blank	0.000	05:20
Subject Sample # 2	0.150	05:20
Breath Volume	1.488	05:20
Air Blank	0.000	05:21
Control Test	0.076	05:22
Air Blank	0.000	05:22
Diagnostics Check	OK	05:22

Data Download Date/Time: 06/06/2018 10:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,J,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: RIVERA,J
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: BARACALDO

DATE/TIME OF TEST: 05/27/2018 04:02:01

Diagnostics Check	OK	03:55	
Air Blank	0.000	03:55	
Control Test	0.080	03:56	
Air Blank	0.000	03:56	
Subject Sample	# 1	03:59	Subject Test Refused
Breath Volume	0.148	03:59	
Air Blank	0.000	04:00	
Control Test	0.081	04:01	
Air Blank		04:01	Sequence Aborted
Air Blank	0.000	04:02	

Data Download Date/Time: 06/06/2018 10:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDWARDS,LANDON,L
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: HOWELL,COLLIN
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: HAYMAN

DATE/TIME OF TEST: 05/09/2018 23:54:04

Diagnostics Check	OK	23:47
Air Blank	0.000	23:47
Control Test	0.081	23:47
Air Blank	0.000	23:48
Subject Sample # 1	0.000	23:49
Breath Volume	3.218	23:49
Air Blank	0.000	23:49
Air Blank	0.000	23:51
Subject Sample # 2	0.000	23:52
Breath Volume	2.667	23:52
Air Blank	0.000	23:53
Control Test	0.080	23:53
Air Blank	0.000	23:53
Diagnostics Check	OK	23:54

Data Download Date/Time: 05/29/2018 17:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAMPE,CORTLAND,B
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: HEATH,JAKEEM
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 21:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: STONE

DATE/TIME OF TEST: 05/15/2018 22:53:12

Diagnostics Check	OK	22:46
Air Blank	0.000	22:46
Control Test	0.080	22:47
Air Blank	0.000	22:47
Subject Sample # 1	0.195	22:48
Breath Volume	2.085	22:48
Air Blank	0.000	22:49
Air Blank	0.000	22:51
Subject Sample # 2	0.203	22:51
Breath Volume	2.167	22:51
Air Blank	0.000	22:52
Control Test	0.080	22:52
Air Blank	0.000	22:53
Diagnostics Check	OK	22:53

Data Download Date/Time: 05/29/2018 17:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,JAMES,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: WARD,JAMES
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 19:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: WARNOCK

DATE/TIME OF TEST: 05/16/2018 20:21:08

Diagnostics Check	OK	20:17	
Air Blank	0.000	20:18	
Control Test	0.081	20:18	
Air Blank	0.000	20:19	
Subject Sample	# 1	20:19	Subject Test Refused
Breath Volume	0.000	20:19	
Air Blank	0.000	20:20	
Control Test	0.081	20:20	
Air Blank	0.000	20:21	
Diagnostics Check	OK	20:21	

Data Download Date/Time: 05/29/2018 17:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,JAMES,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: WARD,JAMES
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: SINGLETON

DATE/TIME OF TEST: 05/19/2018 03:43:23

Diagnostics Check	OK	03:36
Air Blank	0.000	03:37
Control Test	0.081	03:37
Air Blank	0.000	03:37
Subject Sample # 1	0.182	03:38
Breath Volume	1.382	03:38
Air Blank	0.000	03:39
Air Blank	0.000	03:41
Subject Sample # 2	0.193	03:41
Breath Volume	1.902	03:41
Air Blank	0.000	03:42
Control Test	0.080	03:42
Air Blank	0.000	03:43
Diagnostics Check	OK	03:43

Data Download Date/Time: 05/29/2018 17:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAMPE,CORTLAND,B
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: HEATH,J
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: SWAYZER

DATE/TIME OF TEST: 05/19/2018 04:22:35

Diagnostics Check	OK	04:16
Air Blank	0.000	04:16
Control Test	0.078	04:17
Air Blank	0.000	04:17
Subject Sample # 1	0.230	04:17
Breath Volume	2.367	04:17
Air Blank	0.000	04:18
Air Blank	0.000	04:20
Subject Sample # 2	0.229	04:20
Breath Volume	2.019	04:20
Air Blank	0.000	04:21
Control Test	0.080	04:21
Air Blank	0.000	04:22
Diagnostics Check	OK	04:22

Data Download Date/Time: 05/29/2018 17:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDWARDS,LANDON,L
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: EDWARDS,LANDON
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: ALMUTLAQ

DATE/TIME OF TEST: 05/20/2018 04:24:31

Diagnostics Check	OK	04:14
Air Blank	0.000	04:15
Control Test	0.081	04:15
Air Blank	0.000	04:16
Subject Sample # 1	0.070	04:17
Breath Volume	1.140	04:17
Air Blank	0.000	04:18
Air Blank	0.000	04:20
Subject Sample # 2	0.058	04:22
Breath Volume	2.199	04:22
Air Blank	0.000	04:23
Control Test	0.079	04:23
Air Blank	0.000	04:24
Diagnostics Check	OK	04:24

Data Download Date/Time: 05/29/2018 17:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDWARDS,LANDON,L
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: BRYANT,MICHAEL
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 04:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: EWART

DATE/TIME OF TEST: 05/20/2018 05:14:30

Diagnostics Check	OK	05:06
Air Blank	0.000	05:07
Control Test	0.080	05:07
Air Blank	0.000	05:08
Subject Sample # 1	0.140	05:09
Breath Volume	2.945	05:09
Air Blank	0.000	05:10
Air Blank	0.000	05:12
Subject Sample # 2	0.134	05:12
Breath Volume	1.609	05:12
Air Blank	0.000	05:13
Control Test	0.079	05:13
Air Blank	0.000	05:14
Diagnostics Check	OK	05:14

Data Download Date/Time: 05/29/2018 17:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZIGLER,RICKIE,R
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: ZIGLER,RICKIE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 04:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:33

SUBJECT'S LAST NAME: RECUPIDO

DATE/TIME OF TEST: 05/21/2018 05:54:59

Diagnostics Check	OK	05:53	
Air Blank	0.000	05:54	
Control Test	0.080	05:54	
Air Blank		05:54	RFI Detect
Air Blank		05:54	RFI Detect

Data Download Date/Time: 05/29/2018 17:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZIGLER,RICKIE,R
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: ZIGLER,RICKIE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 04:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:33

SUBJECT'S LAST NAME: RECUPIDO

DATE/TIME OF TEST: 05/21/2018 06:04:21

Diagnostics Check	OK	05:58
Air Blank	0.000	05:58
Control Test	0.081	05:58
Air Blank	0.000	05:59
Subject Sample # 1	0.181	05:59
Breath Volume	1.957	05:59
Air Blank	0.000	06:00
Air Blank	0.000	06:02
Subject Sample # 2	0.172	06:02
Breath Volume	1.667	06:02
Air Blank	0.000	06:03
Control Test	0.079	06:03
Air Blank	0.000	06:04
Diagnostics Check	OK	06:04

Data Download Date/Time: 05/29/2018 17:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNNE,JUSTIN,R
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: DUNNE,JUSTIN
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:46

SUBJECT'S LAST NAME: PUGH

DATE/TIME OF TEST: 05/21/2018 23:13:51

Diagnostics Check	OK	23:07
Air Blank	0.000	23:07
Control Test	0.080	23:07
Air Blank	0.000	23:08
Subject Sample # 1	0.210	23:09
Breath Volume	1.890	23:09
Air Blank	0.000	23:09
Air Blank	0.000	23:11
Subject Sample # 2	0.210	23:12
Breath Volume	1.453	23:12
Air Blank	0.000	23:12
Control Test	0.079	23:13
Air Blank	0.000	23:13
Diagnostics Check	OK	23:13

Data Download Date/Time: 05/29/2018 17:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDWARDS,LANDON,L
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: MATERO,DANIEL
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: REEVES

DATE/TIME OF TEST: 05/25/2018 03:31:37

Diagnostics Check	OK	03:24
Air Blank	0.000	03:24
Control Test	0.080	03:24
Air Blank	0.000	03:25
Subject Sample # 1	0.102	03:27
Breath Volume	1.925	03:27
Air Blank	0.000	03:27
Air Blank	0.000	03:29
Subject Sample # 2	0.108	03:29
Breath Volume	1.128	03:29
Air Blank	0.000	03:30
Control Test	0.078	03:31
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time: 05/29/2018 17:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAMPE,CORTLAND,B
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: LAFASO,T
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 05/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 19:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:59

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 05/29/2018 20:27:51

Diagnostics Check	OK	20:19
Air Blank	0.000	20:20
Control Test	0.080	20:20
Air Blank	0.000	20:21
Subject Sample # 1	0.353	20:21
Breath Volume	1.433	20:21
Air Blank	0.000	20:22
Air Blank	0.000	20:24
Subject Sample # 2	0.334	20:25
Breath Volume	1.453	20:25
Air Blank	0.000	20:26
Control Test	0.080	20:27
Air Blank	0.000	20:27
Diagnostics Check	OK	20:27

Data Download Date/Time: 06/27/2018 13:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,JAMES,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: MCARTHY,EVAN
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 05/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: ROUGH

DATE/TIME OF TEST: 05/30/2018 23:39:46

Diagnostics Check	OK	23:33
Air Blank	0.000	23:33
Control Test	0.081	23:33
Air Blank	0.000	23:34
Subject Sample # 1	0.274	23:35
Breath Volume	4.113	23:35
Air Blank	0.000	23:35
Air Blank	0.000	23:37
Subject Sample # 2	0.254	23:38
Breath Volume	2.109	23:38
Air Blank	0.000	23:38
Control Test	0.079	23:39
Air Blank	0.000	23:39
Diagnostics Check	OK	23:39

Data Download Date/Time: 06/27/2018 13:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAZURYK, MICHAEL, S
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: GOTHARD, RONALD
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 04/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A2
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 05/06/2018 00:19:35

Diagnostics Check	OK	00:12
Air Blank	0.000	00:13
Control Test	0.080	00:13
Air Blank	0.000	00:14
Subject Sample # 1	0.146	00:14
Breath Volume	1.925	00:14
Air Blank	0.000	00:15
Air Blank	0.000	00:17
Subject Sample # 2	0.142	00:17
Breath Volume	2.019	00:17
Air Blank	0.000	00:18
Control Test	0.079	00:18
Air Blank	0.000	00:19
Diagnostics Check	OK	00:19

Data Download Date/Time: 05/15/2018 15:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAZURYK, MICHAEL, S
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: MAZURYK, MICHAEL
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 04/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A2
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 05/08/2018 03:28:26

Diagnostics Check	OK	03:21
Air Blank	0.000	03:22
Control Test	0.080	03:22
Air Blank	0.000	03:22
Subject Sample # 1	0.167	03:23
Breath Volume	1.910	03:23
Air Blank	0.000	03:24
Air Blank	0.000	03:26
Subject Sample # 2	0.183	03:26
Breath Volume	2.355	03:26
Air Blank	0.000	03:27
Control Test	0.080	03:27
Air Blank	0.000	03:28
Diagnostics Check	OK	03:28

Data Download Date/Time: 05/15/2018 15:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAETSCH,JOSH,E
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: MORENO,STEPHEN
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 04/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A2
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: JUAREZ SOTO

DATE/TIME OF TEST: 05/14/2018 04:15:31

Diagnostics Check	OK	04:03	
Air Blank	0.000	04:03	
Control Test	0.080	04:03	
Air Blank	0.000	04:04	
Subject Sample	# 1 0.191	04:07	Volume Not Met
Breath Volume	0.019	04:07	
Air Blank	0.000	04:08	
Air Blank	0.000	04:10	
Subject Sample	# 2 0.196	04:10	
Breath Volume	4.203	04:10	
Air Blank	0.000	04:11	
Air Blank	0.000	04:13	
Subject Sample	# 3 0.199	04:13	
Breath Volume	3.238	04:13	
Air Blank	0.000	04:14	
Control Test	0.077	04:14	
Air Blank	0.000	04:15	
Diagnostics Check	OK	04:15	

Data Download Date/Time: 05/15/2018 15:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,J,
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: MORENO,S
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 05/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A2
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: GUEVARA HERNANDEZ

DATE/TIME OF TEST: 05/22/2018 23:18:37

Diagnostics Check	OK	23:10
Air Blank	0.000	23:10
Control Test	0.080	23:10
Air Blank	0.000	23:11
Subject Sample # 1	0.161	23:13
Breath Volume	2.218	23:13
Air Blank	0.000	23:14
Air Blank	0.000	23:16
Subject Sample # 2	0.161	23:16
Breath Volume	1.691	23:16
Air Blank	0.000	23:17
Control Test	0.076	23:17
Air Blank	0.000	23:18
Diagnostics Check	OK	23:18

Data Download Date/Time: 06/13/2018 14:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: FALCON

DATE/TIME OF TEST: 05/09/2018 01:38:42

Diagnostics Check	OK	01:32
Air Blank	0.000	01:33
Control Test	0.079	01:33
Air Blank	0.000	01:33
Subject Sample # 1	0.205	01:34
Breath Volume	2.496	01:34
Air Blank	0.000	01:34
Air Blank	0.000	01:36
Subject Sample # 2	0.211	01:37
Breath Volume	2.929	01:37
Air Blank	0.000	01:37
Control Test	0.078	01:38
Air Blank	0.000	01:38
Diagnostics Check	OK	01:38

Data Download Date/Time: 06/01/2018 11:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 05/09/2018 04:49:44

Diagnostics Check	OK	04:46	
Air Blank	0.000	04:47	
Control Test	0.079	04:47	
Air Blank	0.000	04:48	
Subject Sample	# 1	04:48	Subject Test Refused
Breath Volume	0.000	04:48	
Air Blank	0.000	04:48	
Control Test	0.079	04:49	
Air Blank	0.000	04:49	
Diagnostics Check	OK	04:49	

Data Download Date/Time: 06/01/2018 11:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: BERRY

DATE/TIME OF TEST: 05/10/2018 03:56:56

Diagnostics Check	OK	03:54	
Air Blank	0.000	03:54	
Control Test	0.080	03:54	
Air Blank	0.000	03:55	
Subject Sample	# 1	03:55	Subject Test Refused
Breath Volume	0.000	03:55	
Air Blank	0.000	03:55	
Control Test	0.080	03:56	
Air Blank	0.000	03:56	
Diagnostics Check	OK	03:56	

Data Download Date/Time: 06/01/2018 11:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:56

SUBJECT'S LAST NAME: WORLEY

DATE/TIME OF TEST: 05/11/2018 05:29:10

Diagnostics Check	OK	05:23
Air Blank	0.000	05:23
Control Test	0.080	05:24
Air Blank	0.000	05:24
Subject Sample # 1	0.000	05:24
Breath Volume	1.496	05:24
Air Blank	0.000	05:25
Air Blank	0.000	05:27
Subject Sample # 2	0.000	05:27
Breath Volume	1.507	05:27
Air Blank	0.000	05:28
Control Test	0.079	05:28
Air Blank	0.000	05:29
Diagnostics Check	OK	05:29

Data Download Date/Time: 06/01/2018 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: ROBLES

DATE/TIME OF TEST: 05/12/2018 02:29:04

Diagnostics Check	OK	02:22
Air Blank	0.000	02:23
Control Test	0.079	02:23
Air Blank	0.000	02:23
Subject Sample # 1	0.132	02:24
Breath Volume	1.886	02:24
Air Blank	0.000	02:25
Air Blank	0.000	02:26
Subject Sample # 2	0.131	02:27
Breath Volume	1.878	02:27
Air Blank	0.000	02:28
Control Test	0.079	02:28
Air Blank	0.000	02:28
Diagnostics Check	OK	02:29

Data Download Date/Time: 06/01/2018 11:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: ARGUETA-MERIDA

DATE/TIME OF TEST: 05/13/2018 05:03:05

Diagnostics Check	OK	04:56
Air Blank	0.000	04:57
Control Test	0.079	04:57
Air Blank	0.000	04:58
Subject Sample # 1	0.182	04:58
Breath Volume	4.148	04:58
Air Blank	0.000	04:59
Air Blank	0.000	05:01
Subject Sample # 2	0.176	05:01
Breath Volume	3.085	05:01
Air Blank	0.000	05:02
Control Test	0.078	05:02
Air Blank	0.000	05:02
Diagnostics Check	OK	05:03

Data Download Date/Time: 06/01/2018 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: ANGULO

DATE/TIME OF TEST: 05/17/2018 02:13:49

Diagnostics Check	OK	02:10	
Air Blank	0.000	02:11	
Control Test	0.079	02:11	
Air Blank	0.000	02:12	
Subject Sample	# 1	02:12	Subject Test Refused
Breath Volume	0.000	02:12	
Air Blank	0.000	02:12	
Control Test	0.078	02:13	
Air Blank	0.000	02:13	
Diagnostics Check	OK	02:13	

Data Download Date/Time: 06/01/2018 11:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 05/23/2018 03:10:08

Diagnostics Check	OK	03:07	
Air Blank	0.000	03:07	
Control Test	0.079	03:08	
Air Blank	0.000	03:08	
Subject Sample	# 1	03:08	Subject Test Refused
Breath Volume	0.000	03:08	
Air Blank	0.000	03:09	
Control Test	0.079	03:09	
Air Blank	0.000	03:10	
Diagnostics Check	OK	03:10	

Data Download Date/Time: 06/01/2018 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VALDEZ,MICHAEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: RECINOS-FAJARDO

DATE/TIME OF TEST: 05/27/2018 02:22:41

Diagnostics Check	OK	02:15
Air Blank	0.000	02:16
Control Test	0.079	02:16
Air Blank	0.000	02:17
Subject Sample # 1	0.223	02:18
Breath Volume	2.468	02:18
Air Blank	0.000	02:18
Air Blank	0.000	02:20
Subject Sample # 2	0.207	02:21
Breath Volume	1.980	02:21
Air Blank	0.000	02:21
Control Test	0.078	02:22
Air Blank	0.000	02:22
Diagnostics Check	OK	02:22

Data Download Date/Time: 06/01/2018 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VALDEZ,MICHAEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: RODRIGUEZ-MALDONADO

DATE/TIME OF TEST: 05/28/2018 01:56:59

Diagnostics Check	OK	01:50
Air Blank	0.000	01:50
Control Test	0.079	01:50
Air Blank	0.000	01:51
Subject Sample # 1	0.174	01:52
Breath Volume	1.800	01:52
Air Blank	0.000	01:53
Air Blank	0.000	01:54
Subject Sample # 2	0.183	01:55
Breath Volume	1.269	01:55
Air Blank	0.000	01:56
Control Test	0.078	01:56
Air Blank	0.000	01:56
Diagnostics Check	OK	01:56

Data Download Date/Time: 06/01/2018 11:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: HERRERA

DATE/TIME OF TEST: 05/28/2018 03:27:49

Diagnostics Check	OK	03:21
Air Blank	0.000	03:21
Control Test	0.079	03:22
Air Blank	0.000	03:22
Subject Sample # 1	0.116	03:23
Breath Volume	3.046	03:23
Air Blank	0.000	03:23
Air Blank	0.000	03:25
Subject Sample # 2	0.118	03:26
Breath Volume	3.101	03:26
Air Blank	0.000	03:26
Control Test	0.079	03:27
Air Blank	0.000	03:27
Diagnostics Check	OK	03:27

Data Download Date/Time: 06/01/2018 11:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGHESE, JIBY,
BREATH TEST OPERATOR AGENCY: Aventura PD
ARRESTING OFFICER: RODRIGUEZ, KATINA
ARRESTING OFFICER AGENCY: Aventura PD
LAST AGENCY INSPECTION: 04/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: CALDERON

DATE/TIME OF TEST: 05/04/2018 03:10:05

Diagnostics Check	OK	03:01	
Air Blank	0.000	03:02	
Control Test	0.079	03:02	
Air Blank	0.000	03:02	
Subject Sample	# 1	03:06	No Sample Provided
Breath Volume	0.000	03:06	
Air Blank	0.000	03:06	
Air Blank	0.000	03:08	
Subject Sample	# 2	03:08	Subject Test Refused
Breath Volume	0.000	03:08	
Air Blank	0.000	03:09	
Control Test	0.079	03:09	
Air Blank	0.000	03:10	
Diagnostics Check	OK	03:10	

Data Download Date/Time: 05/15/2018 11:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAVAS,DAVID,E
BREATH TEST OPERATOR AGENCY: Bal Harbour Vil PD
ARRESTING OFFICER: STEPHENSON,TIFANY
ARRESTING OFFICER AGENCY: Sunny Isles Beach PD
LAST AGENCY INSPECTION: 04/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: MASIELLO

DATE/TIME OF TEST: 05/06/2018 05:40:13

Diagnostics Check	OK	05:30	
Air Blank	0.000	05:30	
Control Test	0.079	05:31	
Air Blank	0.000	05:31	
Subject Sample	# 1	05:34	No Sample Provided
Breath Volume	0.000	05:34	
Air Blank	0.000	05:35	
Air Blank	0.000	05:37	
Subject Sample	# 2	05:38	Subject Test Refused
Breath Volume	0.000	05:38	
Air Blank	0.000	05:39	
Control Test	0.078	05:39	
Air Blank	0.000	05:40	
Diagnostics Check	OK	05:40	

Data Download Date/Time: 05/15/2018 11:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NADAL,JOSE,V
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: NADAL,JOSE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: SALCE JR

DATE/TIME OF TEST: 05/23/2018 03:34:42

Diagnostics Check	OK	03:31	
Air Blank	0.000	03:31	
Control Test	0.078	03:31	
Air Blank	0.000	03:32	
Subject Sample	# 1	03:33	Subject Test Refused
Breath Volume	0.000	03:33	
Air Blank	0.000	03:33	
Control Test	0.080	03:34	
Air Blank	0.000	03:34	
Diagnostics Check	OK	03:34	

Data Download Date/Time: 06/13/2018 10:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHRY, HEATHER,
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: ABRAMSON, SHELBY
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 05/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: OTERO RAMIREZ

DATE/TIME OF TEST: 05/31/2018 00:03:30

Diagnostics Check	OK	00:00	
Air Blank	0.000	00:01	
Control Test	0.078	00:01	
Air Blank	0.000	00:01	
Subject Sample	# 1	00:02	Subject Test Refused
Breath Volume	0.000	00:02	
Air Blank	0.000	00:02	
Control Test	0.078	00:02	
Air Blank	0.000	00:03	
Diagnostics Check	OK	00:03	

Data Download Date/Time: 06/13/2018 10:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: RENGIFO

DATE/TIME OF TEST: 05/05/2018 05:27:34

Diagnostics Check	OK	05:21
Air Blank	0.000	05:21
Control Test	0.079	05:21
Air Blank	0.000	05:22
Subject Sample # 1	0.166	05:23
Breath Volume	2.300	05:23
Air Blank	0.000	05:23
Air Blank	0.000	05:25
Subject Sample # 2	0.159	05:25
Breath Volume	2.191	05:25
Air Blank	0.000	05:26
Control Test	0.077	05:26
Air Blank	0.000	05:27
Diagnostics Check	OK	05:27

Data Download Date/Time: 06/01/2018 10:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: NUNEZ-CORTES

DATE/TIME OF TEST: 05/06/2018 02:32:19

Diagnostics Check	OK	02:25
Air Blank	0.000	02:26
Control Test	0.079	02:26
Air Blank	0.000	02:27
Subject Sample # 1	0.167	02:27
Breath Volume	2.304	02:27
Air Blank	0.000	02:28
Air Blank	0.000	02:30
Subject Sample # 2	0.167	02:30
Breath Volume	2.136	02:30
Air Blank	0.000	02:31
Control Test	0.078	02:31
Air Blank	0.000	02:32
Diagnostics Check	OK	02:32

Data Download Date/Time: 06/01/2018 10:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VARGAS,D
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: ARGOTE

DATE/TIME OF TEST: 05/09/2018 02:28:47

Diagnostics Check	OK	02:22
Air Blank	0.000	02:22
Control Test	0.078	02:23
Air Blank	0.000	02:23
Subject Sample # 1	0.114	02:24
Breath Volume	1.800	02:24
Air Blank	0.000	02:24
Air Blank	0.000	02:26
Subject Sample # 2	0.111	02:27
Breath Volume	1.792	02:27
Air Blank	0.000	02:27
Control Test	0.078	02:28
Air Blank	0.000	02:28
Diagnostics Check	OK	02:28

Data Download Date/Time: 06/01/2018 10:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: SOLANO RAMOS

DATE/TIME OF TEST: 05/09/2018 04:11:04

Diagnostics Check	OK	04:02
Air Blank	0.000	04:03
Control Test	0.078	04:03
Air Blank	0.000	04:04
Subject Sample # 1	0.176	04:06
Breath Volume	2.136	04:06
Air Blank	0.000	04:07
Air Blank	0.000	04:09
Subject Sample # 2	0.170	04:09
Breath Volume	1.402	04:09
Air Blank	0.000	04:10
Control Test	0.077	04:10
Air Blank	0.000	04:10
Diagnostics Check	OK	04:11

Data Download Date/Time: 06/01/2018 10:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,R
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: GUTIERREZ

DATE/TIME OF TEST: 05/12/2018 01:44:36

Diagnostics Check	OK	01:38
Air Blank	0.000	01:38
Control Test	0.078	01:39
Air Blank	0.000	01:39
Subject Sample # 1	0.212	01:40
Breath Volume	2.503	01:40
Air Blank	0.000	01:40
Air Blank	0.000	01:42
Subject Sample # 2	0.211	01:42
Breath Volume	2.843	01:42
Air Blank	0.000	01:43
Control Test	0.078	01:43
Air Blank	0.000	01:44
Diagnostics Check	OK	01:44

Data Download Date/Time: 06/01/2018 10:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: GOUDIE

DATE/TIME OF TEST: 05/13/2018 02:05:29

Diagnostics Check	OK	02:02	
Air Blank	0.000	02:03	
Control Test	0.078	02:03	
Air Blank	0.000	02:03	
Subject Sample	# 1	02:04	Subject Test Refused
Breath Volume	0.000	02:04	
Air Blank	0.000	02:04	
Control Test	0.077	02:04	
Air Blank	0.000	02:05	
Diagnostics Check	OK	02:05	

Data Download Date/Time: 06/01/2018 10:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VARGAS,D
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: QUESADA

DATE/TIME OF TEST: 05/30/2018 04:16:23

Diagnostics Check	OK	04:07
Air Blank	0.000	04:08
Control Test	0.078	04:08
Air Blank	0.000	04:08
Subject Sample # 1	0.030	04:11
Breath Volume	2.464	04:11
Air Blank	0.000	04:12
Air Blank	0.000	04:14
Subject Sample # 2	0.030	04:14
Breath Volume	2.000	04:14
Air Blank	0.000	04:15
Control Test	0.077	04:15
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

Data Download Date/Time: 06/01/2018 10:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAYERS,MICHELLE,H
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: SAYERS,MICHELLE
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: MERCADAL RODRIGUEZ

DATE/TIME OF TEST: 05/15/2018 21:49:36

Diagnostics Check	OK	21:42
Air Blank	0.000	21:42
Control Test	0.080	21:42
Air Blank	0.000	21:43
Subject Sample # 1	0.000	21:45
Breath Volume	4.128	21:45
Air Blank	0.000	21:45
Air Blank	0.000	21:47
Subject Sample # 2	0.000	21:48
Breath Volume	3.816	21:48
Air Blank	0.000	21:48
Control Test	0.078	21:48
Air Blank	0.000	21:49
Diagnostics Check	OK	21:49

Data Download Date/Time: 06/05/2018 21:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAYERS,MICHELLE,H
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: SAYERS,MICHELLE
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 06:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: STONE

DATE/TIME OF TEST: 05/20/2018 06:17:17

Diagnostics Check	OK	06:10
Air Blank	0.000	06:10
Control Test	0.080	06:10
Air Blank	0.000	06:11
Subject Sample # 1	0.221	06:12
Breath Volume	4.898	06:12
Air Blank	0.000	06:13
Air Blank	0.000	06:14
Subject Sample # 2	0.224	06:15
Breath Volume	4.957	06:15
Air Blank	0.000	06:16
Control Test	0.078	06:16
Air Blank	0.000	06:17
Diagnostics Check	OK	06:17

Data Download Date/Time: 06/05/2018 21:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MICHELLE,SAYERS,H
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: SAYERS,MICHELLE
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 04:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:56

SUBJECT'S LAST NAME: LATCHMAN

DATE/TIME OF TEST: 05/29/2018 05:30:30

Diagnostics Check	OK	05:24
Air Blank	0.000	05:24
Control Test	0.079	05:24
Air Blank	0.000	05:25
Subject Sample # 1	0.140	05:25
Breath Volume	2.625	05:25
Air Blank	0.000	05:26
Air Blank	0.000	05:28
Subject Sample # 2	0.139	05:28
Breath Volume	1.968	05:28
Air Blank	0.000	05:29
Control Test	0.078	05:29
Air Blank	0.000	05:30
Diagnostics Check	OK	05:30

Data Download Date/Time: 06/05/2018 21:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: COLLARETTI,ANDREW
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 04/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 05/01/2018 23:41:09

Diagnostics Check	OK	23:30
Air Blank	0.000	23:31
Control Test	0.081	23:31
Air Blank	0.000	23:31
Subject Sample # 1	0.186	23:34
Breath Volume	1.394	23:34
Air Blank	0.000	23:35
Air Blank	0.000	23:36
Subject Sample # 2	0.188	23:39
Breath Volume	1.222	23:39
Air Blank	0.000	23:40
Control Test	0.078	23:40
Air Blank	0.000	23:41
Diagnostics Check	OK	23:41

Data Download Date/Time: 05/11/2018 16:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LAWRENCE,TIMOTHY
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 04/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 09:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:57

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 05/02/2018 10:43:15

Diagnostics Check	OK	10:37
Air Blank	0.000	10:37
Control Test	0.081	10:37
Air Blank	0.000	10:38
Subject Sample # 1	0.000	10:38
Breath Volume	1.250	10:38
Air Blank	0.000	10:39
Air Blank	0.000	10:41
Subject Sample # 2	0.000	10:41
Breath Volume	1.812	10:41
Air Blank	0.000	10:42
Control Test	0.081	10:42
Air Blank	0.000	10:43
Diagnostics Check	OK	10:43

Data Download Date/Time: 05/11/2018 16:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: YOCHUM,CRAIG
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 04/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: PILOTTI

DATE/TIME OF TEST: 05/02/2018 22:52:25

Diagnostics Check	OK	22:45
Air Blank	0.000	22:45
Control Test	0.081	22:45
Air Blank	0.000	22:46
Subject Sample # 1	0.241	22:47
Breath Volume	2.449	22:47
Air Blank	0.000	22:48
Air Blank	0.000	22:50
Subject Sample # 2	0.240	22:50
Breath Volume	3.187	22:50
Air Blank	0.000	22:51
Control Test	0.080	22:51
Air Blank	0.000	22:52
Diagnostics Check	OK	22:52

Data Download Date/Time: 05/11/2018 16:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: KYSOR,MITCHELL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 05/03/2018 01:33:52

Diagnostics Check	OK	01:26
Air Blank	0.000	01:27
Control Test	0.081	01:27
Air Blank	0.000	01:28
Subject Sample # 1	0.145	01:29
Breath Volume	2.285	01:29
Air Blank	0.000	01:29
Air Blank	0.000	01:31
Subject Sample # 2	0.142	01:32
Breath Volume	2.339	01:32
Air Blank	0.000	01:32
Control Test	0.078	01:33
Air Blank	0.000	01:33
Diagnostics Check	OK	01:33

Data Download Date/Time: 05/11/2018 16:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MACKEY,JAMES
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 06:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:55

SUBJECT'S LAST NAME: VENTURA

DATE/TIME OF TEST: 05/06/2018 08:27:03

Diagnostics Check	OK	08:19
Air Blank	0.000	08:20
Control Test	0.081	08:20
Air Blank	0.000	08:21
Subject Sample # 1	0.233	08:21
Breath Volume	1.632	08:21
Air Blank	0.000	08:22
Air Blank	0.000	08:24
Subject Sample # 2	0.233	08:25
Breath Volume	1.992	08:25
Air Blank	0.000	08:26
Control Test	0.080	08:26
Air Blank	0.000	08:26
Diagnostics Check	OK	08:27

Data Download Date/Time: 05/11/2018 16:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SNELGROVE,TROY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:53

SUBJECT'S LAST NAME: DIZ ARILLA

DATE/TIME OF TEST: 05/06/2018 22:40:40

Diagnostics Check	OK	22:27	
Air Blank	0.000	22:28	
Control Test	0.081	22:28	
Air Blank	0.000	22:28	
Subject Sample	# 1 0.093	22:30	
Breath Volume	1.316	22:30	
Air Blank	0.000	22:31	
Air Blank	0.000	22:33	
Subject Sample	# 2	22:36	No Sample Provided
Breath Volume	0.000	22:36	
Air Blank	0.000	22:36	
Air Blank	0.000	22:38	
Subject Sample	# 3	22:39	Subject Test Refused
Breath Volume	0.000	22:39	
Air Blank	0.000	22:39	
Control Test	0.081	22:40	
Air Blank	0.000	22:40	
Diagnostics Check	OK	22:40	

Data Download Date/Time: 05/11/2018 16:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,OWEN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LOGAN,RONALD
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 04/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:28

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 05/08/2018 18:05:32

Diagnostics Check	OK	17:53	
Air Blank	0.000	17:54	
Control Test	0.081	17:54	
Air Blank	0.000	17:55	
Subject Sample	# 1 0.169	17:58	Volume Not Met
Breath Volume	0.144	17:58	
Air Blank	0.000	17:58	
Air Blank	0.000	18:00	
Subject Sample	# 2 0.187	18:03	Volume Not Met
Breath Volume	0.136	18:03	
Air Blank	0.000	18:04	
Control Test	0.078	18:04	
Air Blank	0.000	18:05	
Diagnostics Check	OK	18:05	

Data Download Date/Time: 05/11/2018 16:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LOGAN,RONALD
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 04/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:28

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 05/08/2018 18:23:15

Diagnostics Check	OK	18:09	
Air Blank	0.000	18:10	
Control Test	0.078	18:10	
Air Blank	0.000	18:10	
Subject Sample	# 1 0.194	18:14	Volume Not Met
Breath Volume	0.218	18:14	
Air Blank	0.000	18:14	
Air Blank	0.000	18:16	
Subject Sample	# 2 0.208	18:18	
Breath Volume	2.406	18:18	
Air Blank	0.000	18:19	
Air Blank	0.000	18:21	
Subject Sample	# 3	18:21	Subject Test Refused
Breath Volume	0.000	18:21	
Air Blank	0.000	18:22	
Control Test	0.076	18:22	
Air Blank	0.000	18:23	
Diagnostics Check	OK	18:23	

Data Download Date/Time: 05/11/2018 16:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BORROWS,ANDREW
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 04/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:49

SUBJECT'S LAST NAME: PERHOLTZ

DATE/TIME OF TEST: 05/09/2018 19:23:38

Diagnostics Check	OK	19:12
Air Blank	0.000	19:13
Control Test	0.081	19:13
Air Blank	0.000	19:14
Subject Sample # 1	0.214	19:15
Breath Volume	1.355	19:15
Air Blank	0.000	19:16
Air Blank	0.000	19:17
Subject Sample # 2	0.189	19:18
Breath Volume	1.144	19:18
Air Blank	0.000	19:19
Air Blank	0.000	19:21
Subject Sample # 3	0.195	19:21
Breath Volume	1.707	19:21
Air Blank	0.000	19:22
Control Test	0.079	19:23
Air Blank	0.000	19:23
Diagnostics Check	OK	19:23

Data Download Date/Time: 05/11/2018 16:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ROTHENBERG,JUSTIN
ARRESTING OFFICER AGENCY: Palm Beach PD
LAST AGENCY INSPECTION: 04/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: STECKBECK

DATE/TIME OF TEST: 05/11/2018 02:30:16

Diagnostics Check	OK	02:24
Air Blank	0.000	02:24
Control Test	0.080	02:24
Air Blank	0.000	02:25
Subject Sample # 1	0.143	02:25
Breath Volume	2.363	02:25
Air Blank	0.000	02:26
Air Blank	0.000	02:28
Subject Sample # 2	0.142	02:28
Breath Volume	2.007	02:28
Air Blank	0.000	02:29
Control Test	0.081	02:29
Air Blank	0.000	02:30
Diagnostics Check	OK	02:30

Data Download Date/Time: 05/11/2018 16:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006046

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,TABARIE,G
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SULLIVAN,TABARIE
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: DONELL

DATE/TIME OF TEST: 05/25/2018 04:44:35

Diagnostics Check	OK	04:38
Air Blank	0.000	04:38
Control Test	0.077	04:39
Air Blank	0.000	04:39
Subject Sample # 1	0.141	04:40
Breath Volume	2.398	04:40
Air Blank	0.000	04:40
Air Blank	0.000	04:42
Subject Sample # 2	0.146	04:43
Breath Volume	2.390	04:43
Air Blank	0.000	04:43
Control Test	0.077	04:43
Air Blank	0.000	04:44
Diagnostics Check	OK	04:44

Data Download Date/Time: 05/27/2018 17:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006085

INSTRUMENT REGISTERED TO: US COAST GUARD MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELIZAIRE,MARC,K
BREATH TEST OPERATOR AGENCY: USCG Miami
ARRESTING OFFICER: SAMPLE,CARSON
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 04/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: GAZIC

DATE/TIME OF TEST: 05/27/2018 22:53:39

Diagnostics Check	OK	22:50	
Air Blank	0.000	22:50	
Control Test	0.080	22:51	
Air Blank	0.000	22:51	
Subject Sample	# 1	22:52	Subject Test Refused
Breath Volume	0.000	22:52	
Air Blank	0.000	22:52	
Control Test	0.081	22:53	
Air Blank	0.000	22:53	
Diagnostics Check	OK	22:53	

Data Download Date/Time: 05/31/2018 11:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PECCI,Z,A
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: PECCI,Z
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: BAXLEY

DATE/TIME OF TEST: 05/01/2018 03:00:21

Diagnostics Check	OK	02:53
Air Blank	0.000	02:54
Control Test	0.081	02:54
Air Blank	0.000	02:55
Subject Sample # 1	0.067	02:55
Breath Volume	1.605	02:55
Air Blank	0.000	02:56
Air Blank	0.000	02:58
Subject Sample # 2	0.059	02:58
Breath Volume	1.796	02:58
Air Blank	0.000	02:59
Control Test	0.080	02:59
Air Blank	0.000	03:00
Diagnostics Check	OK	03:00

Data Download Date/Time: 05/27/2018 16:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLASZYK,JACOB,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BLASZYK,JACOB
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: BONAN

DATE/TIME OF TEST: 05/02/2018 21:48:14

Diagnostics Check	OK	21:41
Air Blank	0.000	21:42
Control Test	0.080	21:42
Air Blank	0.000	21:43
Subject Sample # 1	0.127	21:43
Breath Volume	3.589	21:43
Air Blank	0.000	21:44
Air Blank	0.000	21:46
Subject Sample # 2	0.127	21:46
Breath Volume	3.582	21:46
Air Blank	0.000	21:47
Control Test	0.080	21:47
Air Blank	0.000	21:48
Diagnostics Check	OK	21:48

Data Download Date/Time: 05/27/2018 16:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENOVIL, GEDEON,
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BRENOVIL, GEDEON
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 05/03/2018 02:35:56

Diagnostics Check	OK	02:28
Air Blank	0.000	02:28
Control Test	0.080	02:29
Air Blank	0.000	02:29
Subject Sample # 1	0.147	02:31
Breath Volume	1.605	02:31
Air Blank	0.000	02:31
Air Blank	0.000	02:33
Subject Sample # 2	0.153	02:34
Breath Volume	1.988	02:34
Air Blank	0.000	02:35
Control Test	0.079	02:35
Air Blank	0.000	02:35
Diagnostics Check	OK	02:35

Data Download Date/Time: 05/27/2018 16:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: COE

DATE/TIME OF TEST: 05/05/2018 01:46:59

Diagnostics Check	OK	01:40
Air Blank	0.000	01:41
Control Test	0.080	01:41
Air Blank	0.000	01:42
Subject Sample # 1	0.163	01:42
Breath Volume	1.800	01:42
Air Blank	0.000	01:43
Air Blank	0.000	01:45
Subject Sample # 2	0.164	01:45
Breath Volume	1.875	01:45
Air Blank	0.000	01:46
Control Test	0.080	01:46
Air Blank	0.000	01:46
Diagnostics Check	OK	01:46

Data Download Date/Time: 05/27/2018 16:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARROLL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: CARROLL,MICHAEL
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: BUI
ARREST TIME: 17:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:27

SUBJECT'S LAST NAME: HERSON

DATE/TIME OF TEST: 05/05/2018 18:52:52

Diagnostics Check	OK	18:47	
Air Blank	0.000	18:48	
Control Test	0.080	18:48	
Air Blank	0.000	18:48	
Subject Sample	# 1	18:51	Subject Test Refused
Breath Volume	0.019	18:51	
Air Blank	0.000	18:51	
Control Test	0.082	18:52	
Air Blank	0.000	18:52	
Diagnostics Check	OK	18:52	

Data Download Date/Time: 05/27/2018 16:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PECCI,Z,A
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: PECCI,Z
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: KLAFFER

DATE/TIME OF TEST: 05/06/2018 01:26:10

Diagnostics Check	OK	01:15
Air Blank	0.000	01:15
Control Test	0.080	01:15
Air Blank	0.000	01:16
Subject Sample # 1	0.191	01:18
Breath Volume	1.183	01:18
Air Blank	0.000	01:19
Air Blank	0.000	01:20
Subject Sample # 2	0.213	01:21
Breath Volume	1.171	01:21
Air Blank	0.000	01:22
Air Blank	0.000	01:23
Subject Sample # 3	0.205	01:24
Breath Volume	1.484	01:24
Air Blank	0.000	01:25
Control Test	0.078	01:25
Air Blank	0.000	01:26
Diagnostics Check	OK	01:26

Data Download Date/Time: 05/27/2018 16:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREGORY,AMANDA,M
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: GREGORY,AMANDA
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: BOTSFORD

DATE/TIME OF TEST: 05/06/2018 23:23:27

Diagnostics Check	OK	23:20	
Air Blank	0.000	23:21	
Control Test	0.080	23:21	
Air Blank	0.000	23:21	
Subject Sample	# 1	23:22	Subject Test Refused
Breath Volume	0.000	23:22	
Air Blank	0.000	23:22	
Control Test	0.079	23:22	
Air Blank	0.000	23:23	
Diagnostics Check	OK	23:23	

Data Download Date/Time: 05/27/2018 16:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: OUELLETTE

DATE/TIME OF TEST: 05/07/2018 00:36:37

Diagnostics Check	OK	00:30
Air Blank	0.000	00:30
Control Test	0.079	00:30
Air Blank	0.000	00:31
Subject Sample # 1	0.169	00:32
Breath Volume	2.929	00:32
Air Blank	0.000	00:32
Air Blank	0.000	00:34
Subject Sample # 2	0.159	00:35
Breath Volume	2.367	00:35
Air Blank	0.000	00:35
Control Test	0.078	00:36
Air Blank	0.000	00:36
Diagnostics Check	OK	00:36

Data Download Date/Time: 05/27/2018 16:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENOVIL,GEDEON,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BRENOVIL,GEDEON
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:27

SUBJECT'S LAST NAME: FINLEY

DATE/TIME OF TEST: 05/08/2018 21:06:08

Diagnostics Check	OK	20:54	
Air Blank	0.000	20:54	
Control Test	0.079	20:54	
Air Blank	0.000	20:55	
Subject Sample	# 1 0.194	20:58	Volume Not Met
Breath Volume	0.097	20:58	
Air Blank	0.000	20:59	
Air Blank	0.000	21:01	
Subject Sample	# 2 0.218	21:04	Volume Not Met
Breath Volume	0.078	21:04	
Air Blank	0.000	21:05	
Control Test	0.079	21:05	
Air Blank	0.000	21:06	
Diagnostics Check	OK	21:06	

Data Download Date/Time: 05/27/2018 16:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENOVIL,GEDEON,
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BRENOVIL,GEDEON
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:27

SUBJECT'S LAST NAME: FINLEY

DATE/TIME OF TEST: 05/08/2018 21:19:30

Diagnostics Check	OK	21:08	
Air Blank	0.000	21:08	
Control Test	0.078	21:08	
Air Blank	0.000	21:09	
Subject Sample	# 1 0.187	21:12	Volume Not Met
Breath Volume	0.417	21:12	
Air Blank	0.000	21:13	
Air Blank	0.000	21:14	
Subject Sample	# 2	21:18	No Sample Provided
Breath Volume	0.000	21:18	
Air Blank	0.000	21:18	
Control Test	0.076	21:18	
Air Blank	0.000	21:19	
Diagnostics Check	OK	21:19	

Data Download Date/Time: 05/27/2018 16:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLASZYK,JACOB,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BLASZYK,JACOB
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: JOHNSON-GRAY

DATE/TIME OF TEST: 05/08/2018 23:29:29

Diagnostics Check	OK	23:13	
Air Blank	0.000	23:13	
Control Test	0.080	23:14	
Air Blank	0.000	23:14	
Subject Sample	# 1 0.167	23:17	Volume Not Met
Breath Volume	1.015	23:17	
Air Blank	0.000	23:18	
Air Blank	0.000	23:20	
Subject Sample	# 2 0.197	23:22	
Breath Volume	1.824	23:22	
Air Blank	0.000	23:23	
Air Blank	0.000	23:24	
Subject Sample	# 3 0.182	23:27	
Breath Volume	1.437	23:27	
Air Blank	0.000	23:28	
Control Test	0.078	23:28	
Air Blank	0.000	23:29	
Diagnostics Check	OK	23:29	

Data Download Date/Time: 05/27/2018 16:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLASZYK,JACOB,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BLASZYK,JACOB
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:22

SUBJECT'S LAST NAME: GODINO

DATE/TIME OF TEST: 05/09/2018 20:50:08

Diagnostics Check	OK	20:43
Air Blank	0.000	20:43
Control Test	0.080	20:44
Air Blank	0.000	20:44
Subject Sample # 1	0.209	20:45
Breath Volume	1.750	20:45
Air Blank	0.000	20:45
Air Blank	0.000	20:47
Subject Sample # 2	0.215	20:48
Breath Volume	1.660	20:48
Air Blank	0.000	20:49
Control Test	0.079	20:49
Air Blank	0.000	20:50
Diagnostics Check	OK	20:50

Data Download Date/Time: 05/27/2018 16:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLASZYK,JACOB,M
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BLASZYK,JACOB
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: WALTON

DATE/TIME OF TEST: 05/11/2018 00:19:59

Diagnostics Check	OK	00:12
Air Blank	0.000	00:13
Control Test	0.080	00:13
Air Blank	0.000	00:14
Subject Sample # 1	0.190	00:14
Breath Volume	2.554	00:14
Air Blank	0.000	00:15
Air Blank	0.000	00:17
Subject Sample # 2	0.179	00:18
Breath Volume	1.472	00:18
Air Blank	0.000	00:19
Control Test	0.080	00:19
Air Blank	0.000	00:19
Diagnostics Check	OK	00:19

Data Download Date/Time: 05/27/2018 16:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: BROUGH

DATE/TIME OF TEST: 05/11/2018 22:57:13

Diagnostics Check	OK	22:50
Air Blank	0.000	22:51
Control Test	0.080	22:51
Air Blank	0.000	22:52
Subject Sample # 1	0.107	22:52
Breath Volume	2.601	22:52
Air Blank	0.000	22:53
Air Blank	0.000	22:55
Subject Sample # 2	0.107	22:55
Breath Volume	2.246	22:55
Air Blank	0.000	22:56
Control Test	0.080	22:56
Air Blank	0.000	22:57
Diagnostics Check	OK	22:57

Data Download Date/Time: 05/27/2018 16:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLIZZI,CHRISTINE,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: POLIZZI,CHRISTINE
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: TRUCCHIO

DATE/TIME OF TEST: 05/14/2018 00:42:41

Diagnostics Check	OK	00:36
Air Blank	0.000	00:36
Control Test	0.080	00:37
Air Blank	0.000	00:37
Subject Sample # 1	0.059	00:38
Breath Volume	1.410	00:38
Air Blank	0.000	00:38
Air Blank	0.000	00:40
Subject Sample # 2	0.057	00:41
Breath Volume	1.378	00:41
Air Blank	0.000	00:41
Control Test	0.080	00:42
Air Blank	0.000	00:42
Diagnostics Check	OK	00:42

Data Download Date/Time: 05/27/2018 16:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PECCI,Z,A
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: PECCI,Z
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: AYERS

DATE/TIME OF TEST: 05/15/2018 02:15:44

Diagnostics Check	OK	02:09
Air Blank	0.000	02:09
Control Test	0.079	02:09
Air Blank	0.000	02:10
Subject Sample # 1	0.113	02:11
Breath Volume	3.035	02:11
Air Blank	0.000	02:11
Air Blank	0.000	02:13
Subject Sample # 2	0.113	02:14
Breath Volume	3.128	02:14
Air Blank	0.000	02:14
Control Test	0.078	02:15
Air Blank	0.000	02:15
Diagnostics Check	OK	02:15

Data Download Date/Time: 05/27/2018 16:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARROLL, MICHAEL, J
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: BREVIK, LONNIE
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: FLEMING

DATE/TIME OF TEST: 05/16/2018 23:08:47

Diagnostics Check	OK	23:05	
Air Blank	0.000	23:06	
Control Test	0.079	23:06	
Air Blank	0.000	23:07	
Subject Sample	# 1	23:07	Subject Test Refused
Breath Volume	0.000	23:07	
Air Blank	0.000	23:07	
Control Test	0.079	23:08	
Air Blank	0.000	23:08	
Diagnostics Check	OK	23:08	

Data Download Date/Time: 05/27/2018 16:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERRA,CORP,
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: SERRA,CORP
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: KENDALL

DATE/TIME OF TEST: 05/17/2018 00:14:33

Diagnostics Check	OK	00:05
Air Blank	0.000	00:05
Control Test	0.079	00:06
Air Blank	0.000	00:06
Subject Sample # 1	0.185	00:07
Breath Volume	2.664	00:07
Air Blank	0.000	00:07
Air Blank	0.000	00:09
Subject Sample # 2	0.162	00:09
Breath Volume	1.351	00:09
Air Blank	0.000	00:10
Air Blank	0.000	00:12
Subject Sample # 3	0.159	00:12
Breath Volume	1.601	00:12
Air Blank	0.000	00:13
Control Test	0.077	00:13
Air Blank	0.000	00:14
Diagnostics Check	OK	00:14

Data Download Date/Time: 05/27/2018 16:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLAGHER,MATTHEW,T
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: GALLAGHER,MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: GARRETT

DATE/TIME OF TEST: 05/17/2018 00:59:18

Diagnostics Check	OK	00:52
Air Blank	0.000	00:53
Control Test	0.079	00:53
Air Blank	0.000	00:54
Subject Sample # 1	0.252	00:54
Breath Volume	1.683	00:54
Air Blank	0.000	00:55
Air Blank	0.000	00:57
Subject Sample # 2	0.248	00:57
Breath Volume	1.578	00:57
Air Blank	0.000	00:58
Control Test	0.079	00:58
Air Blank	0.000	00:59
Diagnostics Check	OK	00:59

Data Download Date/Time: 05/27/2018 16:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PECCI,Z,A
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: PECCI,Z
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: COACHMAN

DATE/TIME OF TEST: 05/19/2018 03:34:22

Diagnostics Check	OK	03:28
Air Blank	0.000	03:28
Control Test	0.080	03:28
Air Blank	0.000	03:29
Subject Sample # 1	0.120	03:29
Breath Volume	1.773	03:29
Air Blank	0.000	03:30
Air Blank	0.000	03:32
Subject Sample # 2	0.120	03:32
Breath Volume	1.484	03:32
Air Blank	0.000	03:33
Control Test	0.080	03:33
Air Blank	0.000	03:34
Diagnostics Check	OK	03:34

Data Download Date/Time: 05/27/2018 16:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: WOLF

DATE/TIME OF TEST: 05/19/2018 18:34:02

Diagnostics Check	OK	18:27
Air Blank	0.000	18:28
Control Test	0.079	18:28
Air Blank	0.000	18:29
Subject Sample # 1	0.175	18:29
Breath Volume	1.921	18:29
Air Blank	0.000	18:30
Air Blank	0.000	18:32
Subject Sample # 2	0.177	18:32
Breath Volume	1.167	18:32
Air Blank	0.000	18:33
Control Test	0.079	18:33
Air Blank	0.000	18:33
Diagnostics Check	OK	18:34

Data Download Date/Time: 05/27/2018 16:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PECCI,Z,A
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: PECCI,Z
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: LIFGREN

DATE/TIME OF TEST: 05/20/2018 02:11:33

Diagnostics Check	OK	02:04
Air Blank	0.000	02:04
Control Test	0.078	02:04
Air Blank	0.000	02:05
Subject Sample # 1	0.109	02:07
Breath Volume	1.304	02:07
Air Blank	0.000	02:07
Air Blank	0.000	02:09
Subject Sample # 2	0.110	02:09
Breath Volume	1.148	02:09
Air Blank	0.000	02:10
Control Test	0.078	02:10
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

Data Download Date/Time: 05/27/2018 16:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLASZYK,JACOB,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BLASZYK,JACOB
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: GASPAR

DATE/TIME OF TEST: 05/21/2018 01:00:13

Diagnostics Check	OK	00:52
Air Blank	0.000	00:52
Control Test	0.080	00:53
Air Blank	0.000	00:53
Subject Sample # 1	0.179	00:55
Breath Volume	1.773	00:55
Air Blank	0.000	00:56
Air Blank	0.000	00:58
Subject Sample # 2	0.187	00:58
Breath Volume	2.277	00:58
Air Blank	0.000	00:59
Control Test	0.079	00:59
Air Blank	0.000	01:00
Diagnostics Check	OK	01:00

Data Download Date/Time: 05/27/2018 16:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLASZYK,JACOB,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BLASZYK,JACOB
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: DIEM

DATE/TIME OF TEST: 05/22/2018 01:33:12

Diagnostics Check	OK	01:26
Air Blank	0.000	01:27
Control Test	0.080	01:27
Air Blank	0.000	01:28
Subject Sample # 1	0.000	01:28
Breath Volume	1.960	01:28
Air Blank	0.000	01:29
Air Blank	0.000	01:31
Subject Sample # 2	0.000	01:31
Breath Volume	2.925	01:31
Air Blank	0.000	01:32
Control Test	0.079	01:32
Air Blank	0.000	01:33
Diagnostics Check	OK	01:33

Data Download Date/Time: 05/27/2018 16:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERRA,CORP,
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: SERRA,CORP
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: PATE

DATE/TIME OF TEST: 05/23/2018 00:25:59

Diagnostics Check	OK	00:19
Air Blank	0.000	00:20
Control Test	0.080	00:20
Air Blank	0.000	00:20
Subject Sample # 1	0.120	00:21
Breath Volume	3.039	00:21
Air Blank	0.000	00:22
Air Blank	0.000	00:23
Subject Sample # 2	0.118	00:24
Breath Volume	2.355	00:24
Air Blank	0.000	00:25
Control Test	0.079	00:25
Air Blank	0.000	00:25
Diagnostics Check	OK	00:25

Data Download Date/Time: 05/27/2018 16:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLASZYK,JACOB,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BLASZYK,JACOB
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: AGUILAR MENDEZ

DATE/TIME OF TEST: 05/27/2018 00:53:13

Diagnostics Check	OK	00:36	
Air Blank	0.000	00:37	
Control Test	0.080	00:37	
Air Blank	0.000	00:38	
Subject Sample	# 1 0.219	00:40	
Breath Volume	1.289	00:40	
Air Blank	0.000	00:41	
Air Blank	0.000	00:42	
Subject Sample	# 2 0.204	00:45	Volume Not Met
Breath Volume	0.132	00:45	
Air Blank	0.000	00:46	
Air Blank	0.000	00:48	
Subject Sample	# 3 0.215	00:51	Volume Not Met
Breath Volume	0.117	00:51	
Air Blank	0.000	00:52	
Control Test	0.077	00:52	
Air Blank	0.000	00:53	
Diagnostics Check	OK	00:53	

Data Download Date/Time: 05/27/2018 16:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENOVIL, GEDEON,
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BRENOVIL, GEDEON
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 05/27/2018 01:36:15

Diagnostics Check	OK	01:29
Air Blank	0.000	01:30
Control Test	0.079	01:30
Air Blank	0.000	01:30
Subject Sample # 1	0.194	01:31
Breath Volume	3.292	01:31
Air Blank	0.000	01:32
Air Blank	0.000	01:34
Subject Sample # 2	0.191	01:34
Breath Volume	3.359	01:34
Air Blank	0.000	01:35
Control Test	0.076	01:35
Air Blank	0.000	01:36
Diagnostics Check	OK	01:36

Data Download Date/Time: 05/27/2018 16:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGERVIL,CLAUDY,P
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: ANGERVIL,CLAUDY
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 05/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:11

SUBJECT'S LAST NAME: PAPE

DATE/TIME OF TEST: 05/28/2018 20:39:11

Diagnostics Check	OK	20:32	
Air Blank	0.000	20:33	
Control Test	0.079	20:33	
Air Blank	0.000	20:34	
Subject Sample	# 1 0.027	20:34	
Breath Volume	1.785	20:34	
Air Blank	0.000	20:35	
Air Blank	0.000	20:37	
Subject Sample	# 2	20:37	Slope Not Met
Breath Volume	2.187	20:37	
Air Blank	0.000	20:38	
Control Test	0.078	20:38	
Air Blank	0.000	20:39	
Diagnostics Check	OK	20:39	

Data Download Date/Time: 06/26/2018 15:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGERVIL,CLAUDY,P
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: ANGERVIL,CLAUDY
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 05/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:11

SUBJECT'S LAST NAME: PAPE

DATE/TIME OF TEST: 05/28/2018 20:48:45

Diagnostics Check	OK	20:42
Air Blank	0.000	20:43
Control Test	0.083	20:43
Air Blank	0.000	20:44
Subject Sample # 1	0.016	20:44
Breath Volume	1.667	20:44
Air Blank	0.000	20:44
Air Blank	0.000	20:46
Subject Sample # 2	0.015	20:47
Breath Volume	1.738	20:47
Air Blank	0.000	20:47
Control Test	0.078	20:48
Air Blank	0.000	20:48
Diagnostics Check	OK	20:48

Data Download Date/Time: 06/26/2018 15:50