

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PATSCH,JOSHUA,K  
BREATH TEST OPERATOR AGENCY: Volusia County SO  
ARRESTING OFFICER: PATSCH,JOSHUA  
ARRESTING OFFICER AGENCY: Volusia County SO  
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 203498  
DRY GAS EXPIRY DATE: 12/10/2018  
VIOLATION CODE: DUI  
ARREST TIME: 21:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: GREENWAY

DATE/TIME OF TEST: 07/03/2018 23:13:04

---

Diagnostics Check		OK	23:01	
Air Blank		0.000	23:01	
Control Test		0.080	23:02	
Air Blank		0.000	23:02	
Subject Sample	# 1	0.135	23:05	Volume Not Met
Breath Volume		0.207	23:05	
Air Blank		0.000	23:06	
Air Blank		0.000	23:08	
Subject Sample	# 2	0.114	23:11	Volume Not Met
Breath Volume		0.128	23:11	
Air Blank		0.000	23:12	
Control Test		0.081	23:12	
Air Blank		0.000	23:12	
Diagnostics Check		OK	23:13	

---

Data Download Date/Time: 07/30/2018 12:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MITCHELL,ROBERT,J  
BREATH TEST OPERATOR AGENCY: Volusia County SO  
ARRESTING OFFICER: LEDOUX,LANCE  
ARRESTING OFFICER AGENCY: Volusia County SO  
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 203498  
DRY GAS EXPIRY DATE: 12/10/2018  
VIOLATION CODE: DUI  
ARREST TIME: 23:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 07/11/2018 23:52:42

---

Diagnostics Check	OK	23:44
Air Blank	0.000	23:45
Control Test	0.080	23:45
Air Blank	0.000	23:45
Subject Sample # 1	0.007	23:47
Breath Volume	1.800	23:47
Air Blank	0.000	23:48
Air Blank	0.000	23:50
Subject Sample # 2	0.000	23:51
Breath Volume	1.445	23:51
Air Blank	0.000	23:51
Control Test	0.079	23:52
Air Blank	0.000	23:52
Diagnostics Check	OK	23:52

---

Data Download Date/Time: 07/30/2018 12:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PATSCH,JOSHUA,K  
BREATH TEST OPERATOR AGENCY: Volusia County SO  
ARRESTING OFFICER: PATSCH,JOSHUA  
ARRESTING OFFICER AGENCY: Volusia County SO  
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 203498  
DRY GAS EXPIRY DATE: 12/10/2018  
VIOLATION CODE: DUI  
ARREST TIME: 22:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 07/15/2018 23:02:38

---

Diagnostics Check	OK	22:52
Air Blank	0.000	22:52
Control Test	0.080	22:53
Air Blank	0.000	22:53
Subject Sample # 1	0.204	22:54
Breath Volume	4.335	22:54
Air Blank	0.000	22:55
Air Blank	0.000	22:57
Subject Sample # 2	0.167	22:58
Breath Volume	1.441	22:58
Air Blank	0.000	22:58
Air Blank	0.000	23:00
Subject Sample # 3	0.187	23:01
Breath Volume	3.671	23:01
Air Blank	0.000	23:01
Control Test	0.079	23:02
Air Blank	0.000	23:02
Diagnostics Check	OK	23:02

---

Data Download Date/Time: 07/30/2018 12:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVIS,ANDREW,J  
BREATH TEST OPERATOR AGENCY: Volusia County SO  
ARRESTING OFFICER: MAXWELL,TIMOTHY  
ARRESTING OFFICER AGENCY: Volusia County SO  
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 203498  
DRY GAS EXPIRY DATE: 12/10/2018  
VIOLATION CODE: DUI  
ARREST TIME: 23:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: LUCAS

DATE/TIME OF TEST: 07/19/2018 00:16:30

---

Diagnostics Check	OK	00:11	
Air Blank	0.000	00:11	
Control Test	0.080	00:11	
Air Blank	0.000	00:12	
Subject Sample	# 1	00:15	Subject Test Refused
Breath Volume	0.257	00:15	
Air Blank	0.000	00:15	
Control Test	0.083	00:15	
Air Blank	0.000	00:16	
Diagnostics Check	OK	00:16	

---

Data Download Date/Time: 07/30/2018 12:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES,BRENT,M  
BREATH TEST OPERATOR AGENCY: Volusia County SO  
ARRESTING OFFICER: JONES,BRENT  
ARRESTING OFFICER AGENCY: Volusia County SO  
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 203498  
DRY GAS EXPIRY DATE: 12/10/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: LEWIS-WARD

DATE/TIME OF TEST: 07/26/2018 03:22:01

---

Diagnostics Check		OK	03:15
Air Blank		0.000	03:16
Control Test		0.080	03:16
Air Blank		0.000	03:16
Subject Sample	# 1	0.151	03:17
Breath Volume		2.800	03:17
Air Blank		0.000	03:18
Air Blank		0.000	03:19
Subject Sample	# 2	0.140	03:20
Breath Volume		2.890	03:20
Air Blank		0.000	03:21
Control Test		0.080	03:21
Air Blank		0.000	03:21
Diagnostics Check		OK	03:22

---

Data Download Date/Time: 07/30/2018 12:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MITCHELL,ROBERT,J  
BREATH TEST OPERATOR AGENCY: Volusia County SO  
ARRESTING OFFICER: JONES,BRENT  
ARRESTING OFFICER AGENCY: Volusia County SO  
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 203498  
DRY GAS EXPIRY DATE: 12/10/2018  
VIOLATION CODE: DUI  
ARREST TIME: 21:31  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: DALEY

DATE/TIME OF TEST: 07/26/2018 22:27:04

---

Diagnostics Check	OK	22:20
Air Blank	0.000	22:21
Control Test	0.080	22:21
Air Blank	0.000	22:22
Subject Sample # 1	0.057	22:22
Breath Volume	1.503	22:22
Air Blank	0.000	22:23
Air Blank	0.000	22:25
Subject Sample # 2	0.055	22:25
Breath Volume	2.035	22:25
Air Blank	0.000	22:26
Control Test	0.080	22:26
Air Blank	0.000	22:26
Diagnostics Check	OK	22:27

---

Data Download Date/Time: 07/30/2018 12:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PATSCH,JOSHUA,K  
BREATH TEST OPERATOR AGENCY: Volusia County SO  
ARRESTING OFFICER: PATSCH,JOSHUA  
ARRESTING OFFICER AGENCY: Volusia County SO  
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 203498  
DRY GAS EXPIRY DATE: 12/10/2018  
VIOLATION CODE: DUI  
ARREST TIME: 00:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: QUEEN

DATE/TIME OF TEST: 07/28/2018 01:22:19

---

Diagnostics Check	OK	01:19	
Air Blank	0.000	01:19	
Control Test	0.080	01:20	
Air Blank	0.000	01:20	
Subject Sample	# 1	01:20	Subject Test Refused
Breath Volume	0.000	01:20	
Air Blank	0.000	01:21	
Control Test	0.080	01:21	
Air Blank	0.000	01:22	
Diagnostics Check	OK	01:22	

---

Data Download Date/Time: 07/30/2018 12:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PATSCH,JOSHUA,K  
BREATH TEST OPERATOR AGENCY: Volusia County SO  
ARRESTING OFFICER: PATSCH,JOSHUA  
ARRESTING OFFICER AGENCY: Volusia County SO  
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 203498  
DRY GAS EXPIRY DATE: 12/10/2018  
VIOLATION CODE: DUI  
ARREST TIME: 23:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: DOMURAD

DATE/TIME OF TEST: 07/30/2018 00:40:58

---

Diagnostics Check	OK	00:27	
Air Blank	0.000	00:28	
Control Test	0.079	00:28	
Air Blank	0.000	00:28	
Subject Sample	# 1	00:32	No Sample Provided
Breath Volume	1.699	00:32	
Air Blank	0.000	00:32	
Air Blank	0.000	00:34	
Subject Sample	# 2	00:35	
Breath Volume	4.292	00:35	
Air Blank	0.000	00:36	
Air Blank	0.000	00:37	
Subject Sample	# 3	00:39	
Breath Volume	4.316	00:39	
Air Blank	0.000	00:40	
Control Test	0.080	00:40	
Air Blank	0.000	00:40	
Diagnostics Check	OK	00:40	

---

Data Download Date/Time: 07/30/2018 12:53



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: MONTGOMERY,KENNETH  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 06/05/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 640976  
DRY GAS EXPIRY DATE: 11/21/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:26  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 07/01/2018 03:17:42

---

Diagnostics Check	OK	03:11
Air Blank	0.000	03:11
Control Test	0.080	03:12
Air Blank	0.000	03:12
Subject Sample # 1	0.142	03:13
Breath Volume	3.230	03:13
Air Blank	0.000	03:13
Air Blank	0.000	03:15
Subject Sample # 2	0.135	03:16
Breath Volume	2.882	03:16
Air Blank	0.000	03:16
Control Test	0.080	03:17
Air Blank	0.000	03:17
Diagnostics Check	OK	03:17

---

Data Download Date/Time: 07/06/2018 14:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DROUIN,CAROLINE,  
BREATH TEST OPERATOR AGENCY: St. Augustine PD  
ARRESTING OFFICER: DROUIN,CAROLINE  
ARRESTING OFFICER AGENCY: St. Augustine PD  
LAST AGENCY INSPECTION: 07/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 640976  
DRY GAS EXPIRY DATE: 11/21/2019  
VIOLATION CODE: DUI  
ARREST TIME: 16:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:35

SUBJECT'S LAST NAME: URIAN

DATE/TIME OF TEST: 07/06/2018 17:30:54

---

Diagnostics Check	OK	17:23
Air Blank	0.000	17:24
Control Test	0.080	17:24
Air Blank	0.000	17:24
Subject Sample # 1	0.142	17:25
Breath Volume	1.410	17:25
Air Blank	0.000	17:26
Air Blank	0.000	17:28
Subject Sample # 2	0.139	17:29
Breath Volume	2.359	17:29
Air Blank	0.000	17:29
Control Test	0.080	17:30
Air Blank	0.000	17:30
Diagnostics Check	OK	17:30

---

Data Download Date/Time: 08/14/2018 14:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BROWN,JOHNATHAN,R  
BREATH TEST OPERATOR AGENCY: St. Augustine PD  
ARRESTING OFFICER: BROWN,JOHNATHAN  
ARRESTING OFFICER AGENCY: St. Augustine PD  
LAST AGENCY INSPECTION: 07/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 640976  
DRY GAS EXPIRY DATE: 11/21/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:26  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: RUTHERFORD

DATE/TIME OF TEST: 07/07/2018 02:22:10

---

Diagnostics Check	OK	02:16
Air Blank	0.000	02:16
Control Test	0.081	02:16
Air Blank	0.000	02:17
Subject Sample # 1	0.108	02:17
Breath Volume	2.824	02:17
Air Blank	0.000	02:18
Air Blank	0.000	02:20
Subject Sample # 2	0.115	02:20
Breath Volume	3.093	02:20
Air Blank	0.000	02:21
Control Test	0.080	02:21
Air Blank	0.000	02:22
Diagnostics Check	OK	02:22

---

Data Download Date/Time: 08/14/2018 14:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RADO,RYAN,R  
BREATH TEST OPERATOR AGENCY: St. Augustine PD  
ARRESTING OFFICER: RADO,RYAN  
ARRESTING OFFICER AGENCY: St. Augustine PD  
LAST AGENCY INSPECTION: 07/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 640976  
DRY GAS EXPIRY DATE: 11/21/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:17  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: LEVY

DATE/TIME OF TEST: 07/07/2018 03:05:11

---

Diagnostics Check	OK	02:58
Air Blank	0.000	02:59
Control Test	0.080	02:59
Air Blank	0.000	03:00
Subject Sample # 1	0.000	03:00
Breath Volume	1.593	03:00
Air Blank	0.000	03:01
Air Blank	0.000	03:03
Subject Sample # 2	0.000	03:03
Breath Volume	1.226	03:03
Air Blank	0.000	03:04
Control Test	0.079	03:04
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

---

Data Download Date/Time: 08/14/2018 14:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RADO,RYAN,R  
BREATH TEST OPERATOR AGENCY: St. Augustine PD  
ARRESTING OFFICER: SHIPP,FRANKIE  
ARRESTING OFFICER AGENCY: St. Augustine PD  
LAST AGENCY INSPECTION: 07/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 640976  
DRY GAS EXPIRY DATE: 11/21/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:36  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: GERLACH

DATE/TIME OF TEST: 07/07/2018 03:48:13

---

Diagnostics Check	OK	03:41
Air Blank	0.000	03:42
Control Test	0.080	03:42
Air Blank	0.000	03:42
Subject Sample # 1	0.057	03:43
Breath Volume	1.542	03:43
Air Blank	0.000	03:44
Air Blank	0.000	03:45
Subject Sample # 2	0.051	03:46
Breath Volume	1.531	03:46
Air Blank	0.000	03:47
Control Test	0.079	03:47
Air Blank	0.000	03:48
Diagnostics Check	OK	03:48

---

Data Download Date/Time: 08/14/2018 14:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HENNESSEY, BRANDON, L  
BREATH TEST OPERATOR AGENCY: St. Augustine Beach  
ARRESTING OFFICER: PRESTON, MIKAYLA  
ARRESTING OFFICER AGENCY: St. Augustine Beach  
LAST AGENCY INSPECTION: 07/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 640976  
DRY GAS EXPIRY DATE: 11/21/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: CARPENTER

DATE/TIME OF TEST: 07/11/2018 23:03:02

---

Diagnostics Check	OK	22:56
Air Blank	0.000	22:56
Control Test	0.079	22:57
Air Blank	0.000	22:57
Subject Sample # 1	0.218	22:58
Breath Volume	1.687	22:58
Air Blank	0.000	22:59
Air Blank	0.000	23:00
Subject Sample # 2	0.220	23:01
Breath Volume	1.480	23:01
Air Blank	0.000	23:02
Control Test	0.079	23:02
Air Blank	0.000	23:02
Diagnostics Check	OK	23:03

---

Data Download Date/Time: 08/14/2018 14:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HANNON,KURT,E  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: TOUBAILI,ABDULRAHMAN  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 07/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 640976  
DRY GAS EXPIRY DATE: 11/21/2019  
VIOLATION CODE: DUI  
ARREST TIME: 05:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:30

SUBJECT'S LAST NAME: PATTERSON

DATE/TIME OF TEST: 07/14/2018 06:57:56

---

Diagnostics Check	OK	06:51
Air Blank	0.000	06:51
Control Test	0.081	06:51
Air Blank	0.000	06:52
Subject Sample # 1	0.191	06:53
Breath Volume	3.195	06:53
Air Blank	0.000	06:53
Air Blank	0.000	06:55
Subject Sample # 2	0.189	06:56
Breath Volume	3.621	06:56
Air Blank	0.000	06:57
Control Test	0.077	06:57
Air Blank	0.000	06:57
Diagnostics Check	OK	06:57

---

Data Download Date/Time: 08/14/2018 14:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HANNON,KURT,E  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: EMINISOR,ADAM  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 07/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 640976  
DRY GAS EXPIRY DATE: 11/21/2019  
VIOLATION CODE: DUI  
ARREST TIME: 06:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:50

SUBJECT'S LAST NAME: PAYNE

DATE/TIME OF TEST: 07/14/2018 07:18:31

---

Diagnostics Check	OK	07:11
Air Blank	0.000	07:11
Control Test	0.079	07:12
Air Blank	0.000	07:12
Subject Sample # 1	0.100	07:13
Breath Volume	3.753	07:13
Air Blank	0.000	07:14
Air Blank	0.000	07:16
Subject Sample # 2	0.100	07:16
Breath Volume	1.562	07:16
Air Blank	0.000	07:17
Control Test	0.077	07:17
Air Blank	0.000	07:18
Diagnostics Check	OK	07:18

---

Data Download Date/Time: 08/14/2018 14:49



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HENNESSEY, BRANDON, L  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: SMALL, RYAN  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 07/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 640976  
DRY GAS EXPIRY DATE: 11/21/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:22  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: ANZOLONE

DATE/TIME OF TEST: 07/14/2018 23:07:26

---

Diagnostics Check	OK	23:01
Air Blank	0.000	23:01
Control Test	0.080	23:01
Air Blank	0.000	23:02
Subject Sample # 1	0.170	23:02
Breath Volume	2.148	23:02
Air Blank	0.000	23:03
Air Blank	0.000	23:05
Subject Sample # 2	0.162	23:05
Breath Volume	2.324	23:05
Air Blank	0.000	23:06
Control Test	0.080	23:06
Air Blank	0.000	23:07
Diagnostics Check	OK	23:07

---

Data Download Date/Time: 08/14/2018 14:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HENNESSEY, BRANDON, L  
BREATH TEST OPERATOR AGENCY: St. Augustine PD  
ARRESTING OFFICER: WRIGHT, BRYAN  
ARRESTING OFFICER AGENCY: St. Augustine Beach  
LAST AGENCY INSPECTION: 07/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 640976  
DRY GAS EXPIRY DATE: 11/21/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:55  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: GRUVER

DATE/TIME OF TEST: 07/15/2018 02:37:25

---

Diagnostics Check	OK	02:31
Air Blank	0.000	02:31
Control Test	0.080	02:31
Air Blank	0.000	02:32
Subject Sample # 1	0.000	02:33
Breath Volume	2.410	02:33
Air Blank	0.000	02:33
Air Blank	0.000	02:35
Subject Sample # 2	0.000	02:36
Breath Volume	2.156	02:36
Air Blank	0.000	02:36
Control Test	0.079	02:36
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

---

Data Download Date/Time: 08/14/2018 14:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LIMA, MICHAEL, A  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: MONTGOMERY, KENNETH  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 07/06/2018

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 640976  
DRY GAS EXPIRY DATE: 11/21/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: FABRA

DATE/TIME OF TEST: 07/15/2018 04:27:08

---

Diagnostics Check	OK	04:20
Air Blank	0.000	04:21
Control Test	0.078	04:21
Air Blank	0.000	04:21
Subject Sample # 1	0.147	04:22
Breath Volume	2.886	04:22
Air Blank	0.000	04:23
Air Blank	0.000	04:25
Subject Sample # 2	0.151	04:25
Breath Volume	3.140	04:25
Air Blank	0.000	04:26
Control Test	0.079	04:26
Air Blank	0.000	04:27
Diagnostics Check	OK	04:27

---

Data Download Date/Time: 08/14/2018 14:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLINE,BRUCE,L  
BREATH TEST OPERATOR AGENCY: St. Augustine Beach  
ARRESTING OFFICER: BRIGGS,JASON  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 07/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 640976  
DRY GAS EXPIRY DATE: 11/21/2019  
VIOLATION CODE: DUI  
ARREST TIME: 15:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:26

SUBJECT'S LAST NAME: FERCHALK

DATE/TIME OF TEST: 07/15/2018 16:58:11

---

Diagnostics Check	OK	16:46	
Air Blank	0.000	16:47	
Control Test	0.080	16:47	
Air Blank	0.000	16:48	
Subject Sample	# 1	16:51	No Sample Provided
Breath Volume	0.000	16:51	
Air Blank	0.000	16:51	
Air Blank	0.000	16:53	
Subject Sample	# 2	16:56	No Sample Provided
Breath Volume	0.000	16:56	
Air Blank	0.000	16:57	
Control Test	0.081	16:57	
Air Blank	0.000	16:58	
Diagnostics Check	OK	16:58	

---

Data Download Date/Time: 08/14/2018 14:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LIMA,MICHAEL,A  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: MONTGOMERY,KENNETH  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 07/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 640976  
DRY GAS EXPIRY DATE: 11/21/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:02  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: MUNAFO

DATE/TIME OF TEST: 07/16/2018 02:56:17

---

Diagnostics Check	OK	02:48
Air Blank	0.000	02:49
Control Test	0.081	02:49
Air Blank	0.000	02:50
Subject Sample # 1	0.164	02:51
Breath Volume	1.328	02:51
Air Blank	0.000	02:52
Air Blank	0.000	02:53
Subject Sample # 2	0.167	02:54
Breath Volume	1.761	02:54
Air Blank	0.000	02:55
Control Test	0.079	02:55
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

---

Data Download Date/Time: 08/14/2018 14:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CORREA,JONATHAN,E  
BREATH TEST OPERATOR AGENCY: St. Augustine PD  
ARRESTING OFFICER: HARTING,MATTHEW  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 07/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 640976  
DRY GAS EXPIRY DATE: 11/21/2019  
VIOLATION CODE: DUI  
ARREST TIME: 04:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: BIESS

DATE/TIME OF TEST: 07/21/2018 05:47:26

---

Diagnostics Check		OK	05:40
Air Blank		0.000	05:41
Control Test		0.080	05:41
Air Blank		0.000	05:42
Subject Sample	# 1	0.220	05:42
Breath Volume		1.457	05:42
Air Blank		0.000	05:43
Air Blank		0.000	05:45
Subject Sample	# 2	0.222	05:45
Breath Volume		1.253	05:45
Air Blank		0.000	05:46
Control Test		0.078	05:46
Air Blank		0.000	05:47
Diagnostics Check		OK	05:47

---

Data Download Date/Time: 08/14/2018 14:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CREPPEL,SEAN,M  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: ROLLINS,TIMOTHY  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 07/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 640976  
DRY GAS EXPIRY DATE: 11/21/2019  
VIOLATION CODE: DUI  
ARREST TIME: 09:53  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 10:00

SUBJECT'S LAST NAME: ORTA

DATE/TIME OF TEST: 07/24/2018 10:59:06

---

Diagnostics Check	OK	10:55	
Air Blank	0.000	10:56	
Control Test	0.080	10:56	
Air Blank	0.000	10:57	
Subject Sample	# 1	10:57	RFI Detect
Breath Volume	0.000	10:57	
Air Blank	0.000	10:58	
Control Test	0.080	10:58	
Air Blank	0.000	10:59	
Diagnostics Check	OK	10:59	

---

Data Download Date/Time: 08/14/2018 14:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CREPPEL,SEAN,M  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: ROLLINS,TIMOTHY  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 07/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 640976  
DRY GAS EXPIRY DATE: 11/21/2019  
VIOLATION CODE: DUI  
ARREST TIME: 09:53  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 10:00

SUBJECT'S LAST NAME: ORTA

DATE/TIME OF TEST: 07/24/2018 11:03:27

---

Diagnostics Check	OK	11:00	
Air Blank	0.000	11:01	
Control Test	0.080	11:01	
Air Blank	0.000	11:01	
Subject Sample	# 1	11:02	Subject Test Refused
Breath Volume	0.000	11:02	
Air Blank	0.000	11:02	
Control Test	0.080	11:02	
Air Blank	0.000	11:03	
Diagnostics Check	OK	11:03	

---

Data Download Date/Time: 08/14/2018 14:49



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WINTERS,ERIN,M  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: WINTERS,ERIN  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 07/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 640976  
DRY GAS EXPIRY DATE: 11/21/2019  
VIOLATION CODE: DUI  
ARREST TIME: 20:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: SNYDER

DATE/TIME OF TEST: 07/24/2018 21:18:35

---

Diagnostics Check	OK	21:10
Air Blank	0.000	21:11
Control Test	0.080	21:11
Air Blank	0.000	21:11
Subject Sample # 1	0.144	21:13
Breath Volume	1.437	21:13
Air Blank	0.000	21:13
Air Blank	0.000	21:15
Subject Sample # 2	0.142	21:17
Breath Volume	1.773	21:17
Air Blank	0.000	21:17
Control Test	0.079	21:17
Air Blank	0.000	21:18
Diagnostics Check	OK	21:18

---

Data Download Date/Time: 08/14/2018 14:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HANNON,KURT,E  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: MULLINS,RICHARD  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 07/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 640976  
DRY GAS EXPIRY DATE: 11/21/2019  
VIOLATION CODE: DUI  
ARREST TIME: 05:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:05

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 07/27/2018 06:32:36

---

Diagnostics Check	OK	06:25
Air Blank	0.000	06:26
Control Test	0.079	06:26
Air Blank	0.000	06:27
Subject Sample # 1	0.192	06:27
Breath Volume	1.187	06:27
Air Blank	0.000	06:28
Air Blank	0.000	06:30
Subject Sample # 2	0.188	06:30
Breath Volume	3.824	06:30
Air Blank	0.000	06:31
Control Test	0.076	06:31
Air Blank	0.000	06:32
Diagnostics Check	OK	06:32

---

Data Download Date/Time: 08/14/2018 14:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WINTERS,ERIN,M  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: BACOLOR,AMANTE  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 07/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 640976  
DRY GAS EXPIRY DATE: 11/21/2019  
VIOLATION CODE: DUI  
ARREST TIME: 11:08  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: HUNTER

DATE/TIME OF TEST: 07/29/2018 00:33:47

---

Diagnostics Check	OK	00:28	
Air Blank	0.000	00:29	
Control Test	0.080	00:29	
Air Blank	0.000	00:30	
Subject Sample	# 1	00:32	RFI Detect
Breath Volume	0.277	00:32	
Air Blank		00:33	RFI Detect
Air Blank	0.000	00:33	

---

Data Download Date/Time: 08/14/2018 14:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WINTERS,ERIN,M  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: BACOLOR,AMANTE  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 07/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 640976  
DRY GAS EXPIRY DATE: 11/21/2019  
VIOLATION CODE: DUI  
ARREST TIME: 11:08  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: HUNTER

DATE/TIME OF TEST: 07/29/2018 00:41:51

---

Diagnostics Check		OK	00:35
Air Blank		0.000	00:35
Control Test		0.078	00:35
Air Blank		0.000	00:36
Subject Sample	# 1	0.216	00:37
Breath Volume		1.250	00:37
Air Blank		0.000	00:38
Air Blank		0.000	00:39
Subject Sample	# 2	0.218	00:40
Breath Volume		1.257	00:40
Air Blank		0.000	00:40
Control Test		0.078	00:41
Air Blank		0.000	00:41
Diagnostics Check		OK	00:41

---

Data Download Date/Time: 08/14/2018 14:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WINTERS,ERIN,M  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: WALLACE,RYAN  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 07/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 640976  
DRY GAS EXPIRY DATE: 11/21/2019  
VIOLATION CODE: DUI  
ARREST TIME: 20:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: GARRELL

DATE/TIME OF TEST: 07/29/2018 21:32:45

---

Diagnostics Check	OK	21:26
Air Blank	0.000	21:26
Control Test	0.079	21:27
Air Blank	0.000	21:27
Subject Sample # 1	0.163	21:28
Breath Volume	2.636	21:28
Air Blank	0.000	21:28
Air Blank	0.000	21:30
Subject Sample # 2	0.144	21:31
Breath Volume	2.312	21:31
Air Blank	0.000	21:31
Control Test	0.078	21:32
Air Blank	0.000	21:32
Diagnostics Check	OK	21:32

---

Data Download Date/Time: 08/14/2018 14:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS, TIFFANY, R  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: WILLIAMS, TIFFANY  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 07/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 640976  
DRY GAS EXPIRY DATE: 11/21/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:41  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: HERBKERSMAN

DATE/TIME OF TEST: 07/30/2018 00:39:40

---

Diagnostics Check	OK	00:33
Air Blank	0.000	00:33
Control Test	0.080	00:34
Air Blank	0.000	00:34
Subject Sample # 1	0.000	00:35
Breath Volume	2.308	00:35
Air Blank	0.000	00:35
Air Blank	0.000	00:37
Subject Sample # 2	0.000	00:38
Breath Volume	2.269	00:38
Air Blank	0.000	00:38
Control Test	0.080	00:39
Air Blank	0.000	00:39
Diagnostics Check	OK	00:39

---

Data Download Date/Time: 08/14/2018 14:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BROWN,JASON,G  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: BROWN,JASON  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 07/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 640976  
DRY GAS EXPIRY DATE: 11/21/2019  
VIOLATION CODE: DUI  
ARREST TIME: 19:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:32

SUBJECT'S LAST NAME: LABNO

DATE/TIME OF TEST: 07/30/2018 21:01:15

---

Diagnostics Check	OK	20:52
Air Blank	0.000	20:53
Control Test	0.080	20:53
Air Blank	0.000	20:53
Subject Sample # 1	0.021	20:56
Breath Volume	1.203	20:56
Air Blank	0.000	20:57
Air Blank	0.000	20:59
Subject Sample # 2	0.022	20:59
Breath Volume	2.222	20:59
Air Blank	0.000	21:00
Control Test	0.081	21:00
Air Blank	0.000	21:01
Diagnostics Check	OK	21:01

---

Data Download Date/Time: 08/14/2018 14:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DANIELS,ALAN,M  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: OWENS,ROGER  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 07/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 640976  
DRY GAS EXPIRY DATE: 11/21/2019  
VIOLATION CODE: DUI  
ARREST TIME: 15:07  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 15:50

SUBJECT'S LAST NAME: BUCKLER

DATE/TIME OF TEST: 07/31/2018 16:23:55

---

Diagnostics Check		OK	16:13
Air Blank		0.000	16:13
Control Test		0.080	16:13
Air Blank		0.000	16:14
Subject Sample	# 1	0.458	16:16
Breath Volume		1.304	16:16
Air Blank		0.000	16:16
Air Blank		0.000	16:18
Subject Sample	# 2	0.436	16:19
Breath Volume		1.546	16:19
Air Blank		0.000	16:19
Air Blank		0.000	16:21
Subject Sample	# 3	0.419	16:22
Breath Volume		1.773	16:22
Air Blank		0.000	16:23
Control Test		0.078	16:23
Air Blank		0.000	16:23
Diagnostics Check		OK	16:23

---

Data Download Date/Time: 08/14/2018 14:49



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ACKERMAN,JUSTIN,T  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: DELUCCA,SAMUEL  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 07/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 640976  
DRY GAS EXPIRY DATE: 11/21/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:57  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: CHAPMAN

DATE/TIME OF TEST: 07/31/2018 23:04:03

---

Diagnostics Check	OK	22:53
Air Blank	0.000	22:53
Control Test	0.079	22:54
Air Blank	0.000	22:54
Subject Sample # 1	0.299	22:56
Breath Volume	1.937	22:56
Air Blank	0.000	22:56
Air Blank	0.000	22:58
Subject Sample # 2	0.259	22:59
Breath Volume	1.257	22:59
Air Blank	0.000	23:00
Air Blank	0.000	23:01
Subject Sample # 3	0.278	23:02
Breath Volume	1.375	23:02
Air Blank	0.000	23:03
Control Test	0.076	23:03
Air Blank	0.000	23:03
Diagnostics Check	OK	23:04

---

Data Download Date/Time: 08/14/2018 14:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001136

INSTRUMENT REGISTERED TO: DELAND PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PATEL,AKHIL,M  
BREATH TEST OPERATOR AGENCY: Deland PD  
ARRESTING OFFICER: WATTS,SHANNON  
ARRESTING OFFICER AGENCY: Deland PD  
LAST AGENCY INSPECTION: 06/26/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 83551  
DRY GAS EXPIRY DATE: 10/02/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 07/07/2018 02:44:48

---

Diagnostics Check	OK	02:37
Air Blank	0.000	02:38
Control Test	0.078	02:38
Air Blank	0.000	02:39
Subject Sample # 1	0.247	02:39
Breath Volume	3.285	02:39
Air Blank	0.000	02:40
Air Blank	0.000	02:42
Subject Sample # 2	0.239	02:43
Breath Volume	2.566	02:43
Air Blank	0.000	02:43
Control Test	0.077	02:44
Air Blank	0.000	02:44
Diagnostics Check	OK	02:44

---

Data Download Date/Time: 07/31/2018 11:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001136

INSTRUMENT REGISTERED TO: DELAND PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PATEL, AKHIL, M  
BREATH TEST OPERATOR AGENCY: Deland PD  
ARRESTING OFFICER: COOPER, WILLIAM  
ARRESTING OFFICER AGENCY: Deland PD  
LAST AGENCY INSPECTION: 06/26/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 83551  
DRY GAS EXPIRY DATE: 10/02/2018  
VIOLATION CODE: DUI  
ARREST TIME: 21:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: BLESDOE

DATE/TIME OF TEST: 07/16/2018 22:13:46

---

Diagnostics Check	OK	22:06
Air Blank	0.000	22:07
Control Test	0.078	22:07
Air Blank	0.000	22:08
Subject Sample # 1	0.000	22:09
Breath Volume	1.511	22:09
Air Blank	0.000	22:09
Air Blank	0.000	22:12
Subject Sample # 2	0.000	22:12
Breath Volume	3.179	22:12
Air Blank	0.000	22:12
Control Test	0.077	22:13
Air Blank	0.000	22:13
Diagnostics Check	OK	22:13

---

Data Download Date/Time: 07/31/2018 11:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001136

INSTRUMENT REGISTERED TO: DELAND PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TAGLE, TONY,  
BREATH TEST OPERATOR AGENCY: Deland PD  
ARRESTING OFFICER: GODWIN, CHALEN  
ARRESTING OFFICER AGENCY: Deland PD  
LAST AGENCY INSPECTION: 06/26/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 83551  
DRY GAS EXPIRY DATE: 10/02/2018  
VIOLATION CODE: DUI  
ARREST TIME: 22:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: SILBERSTEIN

DATE/TIME OF TEST: 07/21/2018 23:06:04

---

Diagnostics Check	OK	23:00
Air Blank	0.000	23:00
Control Test	0.078	23:00
Air Blank	0.000	23:01
Subject Sample # 1	0.000	23:01
Breath Volume	2.835	23:01
Air Blank	0.000	23:02
Air Blank	0.000	23:04
Subject Sample # 2	0.000	23:04
Breath Volume	2.039	23:04
Air Blank	0.000	23:05
Control Test	0.077	23:05
Air Blank	0.000	23:05
Diagnostics Check	OK	23:06

---

Data Download Date/Time: 07/31/2018 11:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH,J,G  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: PURITIS,JEFF  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09117080A1  
DRY GAS EXPIRY DATE: 06/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:18  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: PARISH

DATE/TIME OF TEST: 07/25/2018 00:10:36

---

Diagnostics Check		OK	23:58	
Air Blank		0.000	23:59	
Control Test		0.079	23:59	
Air Blank		0.000	00:00	
Subject Sample	# 1	0.220	00:03	Volume Not Met
Breath Volume		0.332	00:03	
Air Blank		0.000	00:04	
Air Blank		0.000	00:05	
Subject Sample	# 2	0.203	00:08	Volume Not Met
Breath Volume		0.675	00:08	
Air Blank		0.000	00:09	
Control Test		0.078	00:09	
Air Blank		0.000	00:10	
Diagnostics Check		OK	00:10	

---

Data Download Date/Time: 07/30/2018 19:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH,J,G  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: PURITIS,JEFF  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09117080A1  
DRY GAS EXPIRY DATE: 06/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:18  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: PARISH

DATE/TIME OF TEST: 07/25/2018 01:13:24

---

Diagnostics Check	OK	01:07	
Air Blank	0.000	01:07	
Control Test	0.079	01:08	
Air Blank	0.000	01:08	
Subject Sample	# 1	01:11	Subject Test Refused
Breath Volume	0.914	01:11	
Air Blank	0.000	01:12	
Control Test	0.081	01:12	
Air Blank	0.000	01:13	
Diagnostics Check	OK	01:13	

---

Data Download Date/Time: 07/30/2018 19:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEDINA,MARMILLY,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: KELLIHER,MICHAEL  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 796576  
DRY GAS EXPIRY DATE: 06/14/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:41  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: DINNA

DATE/TIME OF TEST: 07/13/2018 23:39:55

---

Diagnostics Check	OK	23:33
Air Blank	0.000	23:34
Control Test	0.082	23:34
Air Blank	0.000	23:35
Subject Sample # 1	0.109	23:35
Breath Volume	1.304	23:35
Air Blank	0.000	23:36
Air Blank	0.000	23:37
Subject Sample # 2	0.105	23:38
Breath Volume	1.503	23:38
Air Blank	0.000	23:39
Control Test	0.082	23:39
Air Blank	0.000	23:39
Diagnostics Check	OK	23:39

---

Data Download Date/Time: 08/29/2018 12:59

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEDINA,MARMILLY,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: LOUGHRAN,KEVIN  
ARRESTING OFFICER AGENCY: OTHER Local  
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 796576  
DRY GAS EXPIRY DATE: 06/14/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: HOLMES

DATE/TIME OF TEST: 07/15/2018 03:33:17

---

Diagnostics Check	OK	03:25
Air Blank	0.000	03:25
Control Test	0.082	03:25
Air Blank	0.000	03:26
Subject Sample # 1	0.102	03:27
Breath Volume	1.492	03:27
Air Blank	0.000	03:28
Air Blank	0.000	03:30
Subject Sample # 2	0.102	03:31
Breath Volume	1.402	03:31
Air Blank	0.000	03:32
Control Test	0.081	03:32
Air Blank	0.000	03:33
Diagnostics Check	OK	03:33

---

Data Download Date/Time: 08/29/2018 12:59



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEDINA,MARMILLY,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: TRABUE,THOMAS  
ARRESTING OFFICER AGENCY: Pembroke Pines PD  
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 796576  
DRY GAS EXPIRY DATE: 06/14/2020  
VIOLATION CODE: DUI  
ARREST TIME: 04:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: SANTOS VENTURA

DATE/TIME OF TEST: 07/15/2018 05:11:34

---

Diagnostics Check	OK	05:04
Air Blank	0.000	05:05
Control Test	0.082	05:05
Air Blank	0.000	05:05
Subject Sample # 1	0.152	05:06
Breath Volume	1.804	05:06
Air Blank	0.000	05:07
Air Blank	0.000	05:09
Subject Sample # 2	0.147	05:10
Breath Volume	1.527	05:10
Air Blank	0.000	05:10
Control Test	0.081	05:10
Air Blank	0.000	05:11
Diagnostics Check	OK	05:11

---

Data Download Date/Time: 08/29/2018 12:59

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEDINA,MARMILLY,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: SERPHOS,STEVEN  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 796576  
DRY GAS EXPIRY DATE: 06/14/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:10  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: VOELKER

DATE/TIME OF TEST: 07/16/2018 00:05:32

---

Diagnostics Check	OK	23:58
Air Blank	0.000	23:58
Control Test	0.083	23:58
Air Blank	0.000	23:59
Subject Sample # 1	0.101	00:00
Breath Volume	1.496	00:00
Air Blank	0.000	00:01
Air Blank	0.000	00:03
Subject Sample # 2	0.105	00:04
Breath Volume	1.480	00:04
Air Blank	0.000	00:04
Control Test	0.081	00:04
Air Blank	0.000	00:05
Diagnostics Check	OK	00:05

---

Data Download Date/Time: 08/29/2018 12:59

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WASHINGTON, YVETTE, D  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: KELLIHER, MICHAEL  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 796576  
DRY GAS EXPIRY DATE: 06/14/2020  
VIOLATION CODE: DUI  
ARREST TIME: 16:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 07/19/2018 17:24:10

---

Diagnostics Check	OK	17:14
Air Blank	0.000	17:14
Control Test	0.082	17:15
Air Blank	0.000	17:15
Subject Sample # 1	0.034	17:17
Breath Volume	1.683	17:17
Air Blank	0.000	17:17
Air Blank	0.000	17:19
Subject Sample # 2	0.034	17:22
Breath Volume	1.648	17:22
Air Blank	0.000	17:23
Control Test	0.082	17:23
Air Blank	0.000	17:24
Diagnostics Check	OK	17:24

---

Data Download Date/Time: 08/29/2018 12:59

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEDINA,MARMILLY,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: VALDES,YANELIS  
ARRESTING OFFICER AGENCY: Hallandale Beach PD  
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 796576  
DRY GAS EXPIRY DATE: 06/14/2020  
VIOLATION CODE: DUI  
ARREST TIME: 13:12  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 14:50

SUBJECT'S LAST NAME: MCWILLIAMS

DATE/TIME OF TEST: 07/20/2018 15:34:05

---

Diagnostics Check	OK	15:27
Air Blank	0.000	15:28
Control Test	0.082	15:28
Air Blank	0.000	15:28
Subject Sample # 1	0.000	15:29
Breath Volume	1.593	15:29
Air Blank	0.000	15:30
Air Blank	0.000	15:32
Subject Sample # 2	0.000	15:32
Breath Volume	1.117	15:32
Air Blank	0.000	15:33
Control Test	0.082	15:33
Air Blank	0.000	15:34
Diagnostics Check	OK	15:34

---

Data Download Date/Time: 08/29/2018 12:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEDINA,MARMILLY,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: CLARKE,ANNE  
ARRESTING OFFICER AGENCY: Margate PD  
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 796576  
DRY GAS EXPIRY DATE: 06/14/2020  
VIOLATION CODE: DUI  
ARREST TIME: 18:36  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: FORTALEZA

DATE/TIME OF TEST: 07/20/2018 20:17:36

---

Diagnostics Check		OK	20:07
Air Blank		0.000	20:08
Control Test		0.082	20:08
Air Blank		0.000	20:08
Subject Sample	# 1	0.303	20:09
Breath Volume		1.468	20:09
Air Blank		0.000	20:10
Air Blank		0.000	20:12
Subject Sample	# 2	0.280	20:13
Breath Volume		1.871	20:13
Air Blank		0.000	20:13
Air Blank		0.000	20:15
Subject Sample	# 3	0.275	20:16
Breath Volume		1.761	20:16
Air Blank		0.000	20:16
Control Test		0.081	20:17
Air Blank		0.000	20:17
Diagnostics Check		OK	20:17

---

Data Download Date/Time: 08/29/2018 12:59

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TILLMAN,SHEILA,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: KEATING,KAYLA  
ARRESTING OFFICER AGENCY: Fort Lauderdale PD  
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 796576  
DRY GAS EXPIRY DATE: 06/14/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:44  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: VELSOR

DATE/TIME OF TEST: 07/21/2018 01:36:12

---

Diagnostics Check	OK	01:29
Air Blank	0.000	01:30
Control Test	0.082	01:30
Air Blank	0.000	01:31
Subject Sample # 1	0.175	01:31
Breath Volume	1.492	01:31
Air Blank	0.000	01:32
Air Blank	0.000	01:34
Subject Sample # 2	0.174	01:34
Breath Volume	2.179	01:34
Air Blank	0.000	01:35
Control Test	0.082	01:35
Air Blank	0.000	01:36
Diagnostics Check	OK	01:36

---

Data Download Date/Time: 08/29/2018 12:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEDINA,MARMILLY,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: KEATING,KAYLA  
ARRESTING OFFICER AGENCY: Fort Lauderdale PD  
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 796576  
DRY GAS EXPIRY DATE: 06/14/2020  
VIOLATION CODE: DUI  
ARREST TIME: 19:37  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: HAWTHORNE

DATE/TIME OF TEST: 07/21/2018 20:33:25

---

Diagnostics Check	OK	20:26
Air Blank	0.000	20:27
Control Test	0.082	20:27
Air Blank	0.000	20:28
Subject Sample # 1	0.215	20:28
Breath Volume	2.218	20:28
Air Blank	0.000	20:29
Air Blank	0.000	20:31
Subject Sample # 2	0.215	20:31
Breath Volume	1.890	20:31
Air Blank	0.000	20:32
Control Test	0.081	20:32
Air Blank	0.000	20:33
Diagnostics Check	OK	20:33

---

Data Download Date/Time: 08/29/2018 12:59

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAZIER,ANAYA,S  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: SOUZA,MICHAEL  
ARRESTING OFFICER AGENCY: OTHER State  
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 796576  
DRY GAS EXPIRY DATE: 06/14/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: DAUCH

DATE/TIME OF TEST: 07/22/2018 03:30:56

---

Diagnostics Check	OK	03:24
Air Blank	0.000	03:24
Control Test	0.082	03:24
Air Blank	0.000	03:25
Subject Sample # 1	0.102	03:25
Breath Volume	3.078	03:25
Air Blank	0.000	03:26
Air Blank	0.000	03:28
Subject Sample # 2	0.106	03:29
Breath Volume	2.753	03:29
Air Blank	0.000	03:30
Control Test	0.082	03:30
Air Blank	0.000	03:30
Diagnostics Check	OK	03:30

---

Data Download Date/Time: 08/29/2018 12:59



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEDINA,MARMILLY,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: HENNESSY,JONATHAN  
ARRESTING OFFICER AGENCY: Sunrise PD  
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 796576  
DRY GAS EXPIRY DATE: 06/14/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 07/23/2018 02:30:41

---

Diagnostics Check		OK	02:24
Air Blank		0.000	02:24
Control Test		0.082	02:24
Air Blank		0.000	02:25
Subject Sample	# 1	0.206	02:26
Breath Volume		1.839	02:26
Air Blank		0.000	02:26
Air Blank		0.000	02:28
Subject Sample	# 2	0.207	02:29
Breath Volume		1.800	02:29
Air Blank		0.000	02:29
Control Test		0.082	02:30
Air Blank		0.000	02:30
Diagnostics Check		OK	02:30

---

Data Download Date/Time: 08/29/2018 12:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEDINA,MARMILLY,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: CHIRINO,YANDY  
ARRESTING OFFICER AGENCY: Hollywood PD  
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 796576  
DRY GAS EXPIRY DATE: 06/14/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: PEREZ GODINEZ

DATE/TIME OF TEST: 07/23/2018 03:27:29

---

Diagnostics Check	OK	03:24	
Air Blank	0.000	03:24	
Control Test	0.081	03:24	
Air Blank	0.000	03:25	
Subject Sample	# 1	03:26	RFI Detect
Breath Volume	0.398	03:26	
Air Blank	0.000	03:26	
Control Test	0.082	03:26	
Air Blank	0.000	03:27	
Diagnostics Check	OK	03:27	

---

Data Download Date/Time: 08/29/2018 12:59

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEDINA,MARMILLY,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: CHIRINO,YANDY  
ARRESTING OFFICER AGENCY: Hollywood PD  
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 796576  
DRY GAS EXPIRY DATE: 06/14/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: PEREZ GODINEZ

DATE/TIME OF TEST: 07/23/2018 03:36:43

---

Diagnostics Check	OK	03:30
Air Blank	0.000	03:30
Control Test	0.081	03:30
Air Blank	0.000	03:31
Subject Sample # 1	0.270	03:32
Breath Volume	2.554	03:32
Air Blank	0.000	03:32
Air Blank	0.000	03:34
Subject Sample # 2	0.262	03:35
Breath Volume	2.398	03:35
Air Blank	0.000	03:35
Control Test	0.080	03:36
Air Blank	0.000	03:36
Diagnostics Check	OK	03:36

---

Data Download Date/Time: 08/29/2018 12:59

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TILLMAN,SHEILA,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: SAPP,JOSHUA  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 796576  
DRY GAS EXPIRY DATE: 06/14/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: RATNAPALA

DATE/TIME OF TEST: 07/26/2018 23:31:23

---

Diagnostics Check	OK	23:23
Air Blank	0.000	23:23
Control Test	0.082	23:24
Air Blank	0.000	23:24
Subject Sample # 1	0.278	23:26
Breath Volume	2.253	23:26
Air Blank	0.000	23:27
Air Blank	0.000	23:29
Subject Sample # 2	0.274	23:29
Breath Volume	2.179	23:29
Air Blank	0.000	23:30
Control Test	0.081	23:30
Air Blank	0.000	23:31
Diagnostics Check	OK	23:31

---

Data Download Date/Time: 08/29/2018 12:59

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TILLMAN,SHEILA,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: CARTER,JAMES  
ARRESTING OFFICER AGENCY: Fort Lauderdale PD  
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 796576  
DRY GAS EXPIRY DATE: 06/14/2020  
VIOLATION CODE: DUI  
ARREST TIME: 04:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:12

SUBJECT'S LAST NAME: DESHOMMES

DATE/TIME OF TEST: 07/27/2018 05:41:47

---

Diagnostics Check	OK	05:35
Air Blank	0.000	05:35
Control Test	0.082	05:36
Air Blank	0.000	05:36
Subject Sample # 1	0.127	05:37
Breath Volume	2.328	05:37
Air Blank	0.000	05:37
Air Blank	0.000	05:39
Subject Sample # 2	0.123	05:40
Breath Volume	2.789	05:40
Air Blank	0.000	05:40
Control Test	0.082	05:41
Air Blank	0.000	05:41
Diagnostics Check	OK	05:41

---

Data Download Date/Time: 08/29/2018 12:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAZIER,ANAYA,S  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: ARMSTRONG,BRITTANY  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 796576  
DRY GAS EXPIRY DATE: 06/14/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: BAKALOV

DATE/TIME OF TEST: 07/29/2018 02:09:22

---

Diagnostics Check		OK	01:52	
Air Blank		0.000	01:53	
Control Test		0.081	01:53	
Air Blank		0.000	01:54	
Subject Sample	# 1	0.160	01:57	Volume Not Met
Breath Volume		0.136	01:57	
Air Blank		0.000	01:57	
Air Blank		0.000	01:59	
Subject Sample	# 2	0.173	02:02	
Breath Volume		1.531	02:02	
Air Blank		0.000	02:03	
Air Blank		0.000	02:05	
Subject Sample	# 3	0.171	02:07	
Breath Volume		1.339	02:07	
Air Blank		0.000	02:08	
Control Test		0.081	02:08	
Air Blank		0.000	02:09	
Diagnostics Check		OK	02:09	

---

Data Download Date/Time: 08/29/2018 12:59

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SAPP,JOSHUA,R  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: SAPP,JOSHUA  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 796576  
DRY GAS EXPIRY DATE: 06/14/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: WORMLEY

DATE/TIME OF TEST: 07/29/2018 02:25:15

---

Diagnostics Check	OK	02:19
Air Blank	0.000	02:19
Control Test	0.081	02:19
Air Blank	0.000	02:20
Subject Sample # 1	0.175	02:20
Breath Volume	1.746	02:20
Air Blank	0.000	02:21
Air Blank	0.000	02:23
Subject Sample # 2	0.176	02:23
Breath Volume	1.773	02:23
Air Blank	0.000	02:24
Control Test	0.082	02:24
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

---

Data Download Date/Time: 08/29/2018 12:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAZIER,ANAYA,S  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: BERMAN,MARC  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 796576  
DRY GAS EXPIRY DATE: 06/14/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: VALDES

DATE/TIME OF TEST: 07/30/2018 00:23:36

---

Diagnostics Check	OK	00:17
Air Blank	0.000	00:18
Control Test	0.082	00:18
Air Blank	0.000	00:18
Subject Sample # 1	0.072	00:19
Breath Volume	3.074	00:19
Air Blank	0.000	00:19
Air Blank	0.000	00:21
Subject Sample # 2	0.071	00:22
Breath Volume	3.035	00:22
Air Blank	0.000	00:22
Control Test	0.081	00:23
Air Blank	0.000	00:23
Diagnostics Check	OK	00:23

---

Data Download Date/Time: 08/29/2018 12:59



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAZIER,ANAYA,S  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: CARVALHO,S  
ARRESTING OFFICER AGENCY: Seminole Dept of LE  
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 796576  
DRY GAS EXPIRY DATE: 06/14/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:57  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: QUESADA REYES

DATE/TIME OF TEST: 07/30/2018 02:24:42

---

Diagnostics Check	OK	02:17
Air Blank	0.000	02:18
Control Test	0.082	02:18
Air Blank	0.000	02:19
Subject Sample # 1	0.242	02:20
Breath Volume	1.347	02:20
Air Blank	0.000	02:20
Air Blank	0.000	02:22
Subject Sample # 2	0.235	02:23
Breath Volume	1.105	02:23
Air Blank	0.000	02:23
Control Test	0.081	02:24
Air Blank	0.000	02:24
Diagnostics Check	OK	02:24

---

Data Download Date/Time: 08/29/2018 12:59

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAZIER,ANAYA,S  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: SPEZIALE,ADAM  
ARRESTING OFFICER AGENCY: Margate PD  
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 796576  
DRY GAS EXPIRY DATE: 06/14/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: HERNANDEZ VASQUEZ

DATE/TIME OF TEST: 07/30/2018 03:18:56

---

Diagnostics Check		OK	03:12
Air Blank		0.000	03:13
Control Test		0.081	03:13
Air Blank		0.000	03:14
Subject Sample	# 1	0.262	03:14
Breath Volume		1.875	03:14
Air Blank		0.000	03:15
Air Blank		0.000	03:17
Subject Sample	# 2	0.264	03:17
Breath Volume		3.027	03:17
Air Blank		0.000	03:18
Control Test		0.082	03:18
Air Blank		0.000	03:18
Diagnostics Check		OK	03:18

---

Data Download Date/Time: 08/29/2018 12:59

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SPURLOCK,ROBERT,V  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: SPURLOCK,ROBERT  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 867139  
DRY GAS EXPIRY DATE: 11/15/2020  
VIOLATION CODE: DUI  
ARREST TIME: 15:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:30

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 07/05/2018 18:07:36

---

Diagnostics Check	OK	18:01
Air Blank	0.000	18:02
Control Test	0.078	18:02
Air Blank	0.000	18:03
Subject Sample # 1	0.000	18:03
Breath Volume	1.718	18:03
Air Blank	0.000	18:03
Air Blank	0.000	18:05
Subject Sample # 2	0.000	18:06
Breath Volume	1.519	18:06
Air Blank	0.000	18:06
Control Test	0.081	18:06
Air Blank	0.000	18:07
Diagnostics Check	OK	18:07

---

Data Download Date/Time: 07/24/2018 09:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001152

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PAGE,BRIAN,R  
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD  
ARRESTING OFFICER: PAGE,BRIAN  
ARRESTING OFFICER AGENCY: New Smyrna Beach PD  
LAST AGENCY INSPECTION: 06/23/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A2  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 16:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 07/10/2018 17:37:30

---

Diagnostics Check	Fail	17:37	Voltage/Current Test
Air Blank	0.000	17:37	

---

Data Download Date/Time: 07/21/2018 08:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001152

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PAGE,BRIAN,P  
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD  
ARRESTING OFFICER: SHAW,KYLEE  
ARRESTING OFFICER AGENCY: New Smyrna Beach PD  
LAST AGENCY INSPECTION: 07/21/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A2  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 16:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:49

SUBJECT'S LAST NAME: CHAPPEL

DATE/TIME OF TEST: 07/28/2018 17:19:55

---

Diagnostics Check	OK	17:09	
Air Blank	0.000	17:09	
Control Test	0.079	17:10	
Air Blank	0.000	17:10	
Subject Sample	# 1 0.194	17:13	Volume Not Met
Breath Volume	0.308	17:13	
Air Blank	0.000	17:14	
Air Blank	0.000	17:16	
Subject Sample	# 2	17:18	Subject Test Refused
Breath Volume	0.226	17:18	
Air Blank	0.000	17:18	
Control Test	0.078	17:19	
Air Blank	0.000	17:19	
Diagnostics Check	OK	17:19	

---

Data Download Date/Time: 09/03/2018 07:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CORBIN,STEPHEN,A  
BREATH TEST OPERATOR AGENCY: Volusia County SO  
ARRESTING OFFICER: CORBIN,STEPHEN  
ARRESTING OFFICER AGENCY: Volusia County SO  
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 644476  
DRY GAS EXPIRY DATE: 12/01/2019  
VIOLATION CODE: DUI  
ARREST TIME: 16:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:54

SUBJECT'S LAST NAME: ROUGH

DATE/TIME OF TEST: 07/10/2018 17:30:39

---

Diagnostics Check	OK	17:24
Air Blank	0.000	17:24
Control Test	0.079	17:24
Air Blank	0.000	17:25
Subject Sample # 1	0.330	17:25
Breath Volume	2.335	17:25
Air Blank	0.000	17:26
Air Blank	0.000	17:28
Subject Sample # 2	0.338	17:29
Breath Volume	1.972	17:29
Air Blank	0.000	17:29
Control Test	0.080	17:30
Air Blank	0.000	17:30
Diagnostics Check	OK	17:30

---

Data Download Date/Time: 07/30/2018 12:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WARMINGTON, WILLIAM, G  
BREATH TEST OPERATOR AGENCY: Ormond Beach PD  
ARRESTING OFFICER: WARMINGTON, WILLIAM  
ARRESTING OFFICER AGENCY: Ormond Beach PD  
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 644476  
DRY GAS EXPIRY DATE: 12/01/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: PINAR

DATE/TIME OF TEST: 07/20/2018 00:37:22

---

Diagnostics Check	OK	00:30
Air Blank	0.000	00:31
Control Test	0.081	00:31
Air Blank	0.000	00:32
Subject Sample # 1	0.072	00:32
Breath Volume	2.261	00:32
Air Blank	0.000	00:33
Air Blank	0.000	00:35
Subject Sample # 2	0.074	00:35
Breath Volume	2.136	00:35
Air Blank	0.000	00:36
Control Test	0.082	00:36
Air Blank	0.000	00:37
Diagnostics Check	OK	00:37

---

Data Download Date/Time: 07/30/2018 12:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WARMINGTON, WILLIAM, G  
BREATH TEST OPERATOR AGENCY: Ormond Beach PD  
ARRESTING OFFICER: GARCIA, THOMAS  
ARRESTING OFFICER AGENCY: Ormond Beach PD  
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 644476  
DRY GAS EXPIRY DATE: 12/01/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: PAPPA ABARBANELL

DATE/TIME OF TEST: 07/20/2018 03:09:23

---

Diagnostics Check	OK	03:06	
Air Blank	0.000	03:06	
Control Test	0.081	03:07	
Air Blank	0.000	03:07	
Subject Sample	# 1	03:08	Subject Test Refused
Breath Volume	0.000	03:08	
Air Blank	0.000	03:08	
Control Test	0.079	03:08	
Air Blank	0.000	03:09	
Diagnostics Check	OK	03:09	

---

Data Download Date/Time: 07/30/2018 12:36



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOSS,JEREMY,A  
BREATH TEST OPERATOR AGENCY: Ormond Beach PD  
ARRESTING OFFICER: RAIHALA,JON  
ARRESTING OFFICER AGENCY: Ormond Beach PD  
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 644476  
DRY GAS EXPIRY DATE: 12/01/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: WALTON

DATE/TIME OF TEST: 07/22/2018 00:56:38

---

Diagnostics Check	OK	00:49
Air Blank	0.000	00:49
Control Test	0.082	00:49
Air Blank	0.000	00:50
Subject Sample # 1	0.094	00:52
Breath Volume	1.570	00:52
Air Blank	0.000	00:52
Air Blank	0.000	00:54
Subject Sample # 2	0.098	00:55
Breath Volume	1.929	00:55
Air Blank	0.000	00:55
Control Test	0.082	00:56
Air Blank	0.000	00:56
Diagnostics Check	OK	00:56

---

Data Download Date/Time: 07/30/2018 12:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RUBIN,RICHARD,D  
BREATH TEST OPERATOR AGENCY: Ormond Beach PD  
ARRESTING OFFICER: RAIHALA,JON  
ARRESTING OFFICER AGENCY: Ormond Beach PD  
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 644476  
DRY GAS EXPIRY DATE: 12/01/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: WOOD

DATE/TIME OF TEST: 07/23/2018 02:55:24

---

Diagnostics Check	OK	02:52	
Air Blank	0.000	02:53	
Control Test	0.081	02:53	
Air Blank	0.000	02:53	
Subject Sample	# 1	02:54	Subject Test Refused
Breath Volume	0.000	02:54	
Air Blank	0.000	02:54	
Control Test	0.080	02:54	
Air Blank	0.000	02:55	
Diagnostics Check	OK	02:55	

---

Data Download Date/Time: 07/30/2018 12:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WARMINGTON, WILLIAM, G  
BREATH TEST OPERATOR AGENCY: Ormond Beach PD  
ARRESTING OFFICER: RAIHALA, JON  
ARRESTING OFFICER AGENCY: Ormond Beach PD  
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 644476  
DRY GAS EXPIRY DATE: 12/01/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: BARBACCI

DATE/TIME OF TEST: 07/31/2018 03:54:42

---

Diagnostics Check	OK	03:51	
Air Blank	0.000	03:52	
Control Test	0.081	03:52	
Air Blank	0.000	03:53	
Subject Sample	# 1	03:53	Subject Test Refused
Breath Volume	0.000	03:53	
Air Blank	0.000	03:53	
Control Test	0.080	03:54	
Air Blank	0.000	03:54	
Diagnostics Check	OK	03:54	

---

Data Download Date/Time: 08/10/2018 12:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RUBIN,RICHARD,D  
BREATH TEST OPERATOR AGENCY: Ormond Beach PD  
ARRESTING OFFICER: RAIHALA,JON  
ARRESTING OFFICER AGENCY: Ormond Beach PD  
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814183  
DRY GAS EXPIRY DATE: 07/26/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: SAUM

DATE/TIME OF TEST: 07/07/2018 00:25:04

---

Diagnostics Check		OK	00:18
Air Blank		0.000	00:18
Control Test		0.081	00:18
Air Blank		0.000	00:19
Subject Sample	# 1	0.087	00:20
Breath Volume		1.238	00:20
Air Blank		0.000	00:20
Air Blank		0.000	00:22
Subject Sample	# 2	0.087	00:23
Breath Volume		1.109	00:23
Air Blank		0.000	00:24
Control Test		0.081	00:24
Air Blank		0.000	00:24
Diagnostics Check		OK	00:25

---

Data Download Date/Time: 07/17/2018 10:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RUBIN,RICHARD,D  
BREATH TEST OPERATOR AGENCY: Ormond Beach PD  
ARRESTING OFFICER: RAIHALA,JON  
ARRESTING OFFICER AGENCY: Ormond Beach PD  
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814183  
DRY GAS EXPIRY DATE: 07/26/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: SAUM

DATE/TIME OF TEST: 07/07/2018 00:52:24

---

Diagnostics Check	OK	00:46
Air Blank	0.000	00:46
Control Test	0.081	00:46
Air Blank	0.000	00:47
Subject Sample # 1	0.081	00:47
Breath Volume	1.476	00:47
Air Blank	0.000	00:48
Air Blank	0.000	00:50
Subject Sample # 2	0.074	00:50
Breath Volume	1.707	00:50
Air Blank	0.000	00:51
Control Test	0.081	00:51
Air Blank	0.000	00:52
Diagnostics Check	OK	00:52

---

Data Download Date/Time: 07/17/2018 10:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WARMINGTON, WILLIAM, G  
BREATH TEST OPERATOR AGENCY: Ormond Beach PD  
ARRESTING OFFICER: DEVANE, CIARA  
ARRESTING OFFICER AGENCY: Ormond Beach PD  
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814183  
DRY GAS EXPIRY DATE: 07/26/2020  
VIOLATION CODE: DUI  
ARREST TIME: 12:41  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 15:03

SUBJECT'S LAST NAME: BITTLE

DATE/TIME OF TEST: 07/07/2018 15:32:29

---

Diagnostics Check	OK	15:23
Air Blank	0.000	15:24
Control Test	0.081	15:24
Air Blank	0.000	15:24
Subject Sample # 1	0.000	15:26
Breath Volume	1.289	15:26
Air Blank	0.000	15:26
Air Blank	0.000	15:28
Subject Sample # 2	0.000	15:31
Breath Volume	1.476	15:31
Air Blank	0.000	15:31
Control Test	0.082	15:31
Air Blank	0.000	15:32
Diagnostics Check	OK	15:32

---

Data Download Date/Time: 07/17/2018 10:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001158

INSTRUMENT REGISTERED TO: PONCE INLET PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRAIG,BRADLEY,A  
BREATH TEST OPERATOR AGENCY: Ponce Inlet PD  
ARRESTING OFFICER: CRAIG,BRADLEY  
ARRESTING OFFICER AGENCY: Ponce Inlet PD  
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814183  
DRY GAS EXPIRY DATE: 07/26/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: TROTTER

DATE/TIME OF TEST: 07/09/2018 22:28:00

---

Diagnostics Check	OK	22:24	
Air Blank	0.000	22:25	
Control Test	0.081	22:25	
Air Blank	0.000	22:26	
Subject Sample	# 1	22:26	Subject Test Refused
Breath Volume	0.000	22:26	
Air Blank	0.000	22:27	
Control Test	0.081	22:27	
Air Blank	0.000	22:27	
Diagnostics Check	OK	22:28	

---

Data Download Date/Time: 07/22/2018 06:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001159

INSTRUMENT REGISTERED TO: ORANGE CITY P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KIBBE,BRIAN,J  
BREATH TEST OPERATOR AGENCY: Orange City PD  
ARRESTING OFFICER: JONES,BRENT  
ARRESTING OFFICER AGENCY: Volusia County SO  
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17316080A2  
DRY GAS EXPIRY DATE: 08/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 00:43  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 07/15/2018 01:25:20

---

Diagnostics Check	OK	01:14
Air Blank	0.000	01:14
Control Test	0.081	01:15
Air Blank	0.000	01:15
Subject Sample # 1	0.178	01:19
Breath Volume	1.488	01:19
Air Blank	0.000	01:19
Air Blank	0.000	01:21
Subject Sample # 2	0.165	01:23
Breath Volume	2.734	01:23
Air Blank	0.000	01:24
Control Test	0.081	01:24
Air Blank	0.000	01:25
Diagnostics Check	OK	01:25

---

Data Download Date/Time: 07/15/2018 06:36



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001159

INSTRUMENT REGISTERED TO: ORANGE CITY P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRALSON,SAMUEL,E  
BREATH TEST OPERATOR AGENCY: Orange City PD  
ARRESTING OFFICER: HARRALSON,SAMUEL  
ARRESTING OFFICER AGENCY: Orange City PD  
LAST AGENCY INSPECTION: 07/16/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17316080A2  
DRY GAS EXPIRY DATE: 08/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: PATTERSON

DATE/TIME OF TEST: 07/18/2018 02:02:34

---

Diagnostics Check	OK	01:56
Air Blank	0.000	01:56
Control Test	0.081	01:57
Air Blank	0.000	01:57
Subject Sample # 1	0.174	01:58
Breath Volume	2.183	01:58
Air Blank	0.000	01:58
Air Blank	0.000	02:00
Subject Sample # 2	0.194	02:01
Breath Volume	2.765	02:01
Air Blank	0.000	02:01
Control Test	0.080	02:01
Air Blank	0.000	02:02
Diagnostics Check	OK	02:02

---

Data Download Date/Time: 08/18/2018 06:57

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

INSTRUMENT REGISTERED TO: LADY LAKE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRONK,A,  
BREATH TEST OPERATOR AGENCY: Lady Lake PD  
ARRESTING OFFICER: ALVAREZ,D  
ARRESTING OFFICER AGENCY: Lady Lake PD  
LAST AGENCY INSPECTION: 07/09/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17817080A3  
DRY GAS EXPIRY DATE: 08/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 14:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:17

SUBJECT'S LAST NAME: SIMO

DATE/TIME OF TEST: 07/28/2018 15:02:12

---

Diagnostics Check	OK	14:55
Air Blank	0.000	14:56
Control Test	0.078	14:56
Air Blank	0.000	14:57
Subject Sample # 1	0.059	14:57
Breath Volume	3.175	14:57
Air Blank	0.000	14:58
Air Blank	0.000	15:00
Subject Sample # 2	0.054	15:00
Breath Volume	4.312	15:00
Air Blank	0.000	15:01
Control Test	0.079	15:01
Air Blank	0.000	15:02
Diagnostics Check	OK	15:02

---

Data Download Date/Time: 08/22/2018 15:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

INSTRUMENT REGISTERED TO: LADY LAKE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DURYEAMATTHEW,  
BREATH TEST OPERATOR AGENCY: Lady Lake PD  
ARRESTING OFFICER: DURYEAMATTHEW  
ARRESTING OFFICER AGENCY: Lady Lake PD  
LAST AGENCY INSPECTION: 07/09/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17817080A3  
DRY GAS EXPIRY DATE: 08/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: WAGNER

DATE/TIME OF TEST: 07/29/2018 00:12:57

---

Diagnostics Check	OK	00:10	
Air Blank	0.000	00:10	
Control Test	0.078	00:10	
Air Blank	0.000	00:11	
Subject Sample	# 1	00:11	Subject Test Refused
Breath Volume	0.000	00:11	
Air Blank	0.000	00:12	
Control Test	0.079	00:12	
Air Blank	0.000	00:12	
Diagnostics Check	OK	00:12	

---

Data Download Date/Time: 08/22/2018 15:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001179

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COSCULLUELA,BRYAN,M  
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)  
ARRESTING OFFICER: SAAVEDRA,OSCAR  
ARRESTING OFFICER AGENCY: FFWCC (West Palm)  
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/24/2019  
VIOLATION CODE: BUI  
ARREST TIME: 00:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: JAIME PI

DATE/TIME OF TEST: 07/01/2018 00:19:51

---

Diagnostics Check	OK	00:13
Air Blank	0.000	00:13
Control Test	0.079	00:14
Air Blank	0.000	00:14
Subject Sample # 1	0.152	00:15
Breath Volume	1.824	00:15
Air Blank	0.000	00:15
Air Blank	0.000	00:17
Subject Sample # 2	0.149	00:18
Breath Volume	1.644	00:18
Air Blank	0.000	00:18
Control Test	0.079	00:19
Air Blank	0.000	00:19
Diagnostics Check	OK	00:19

---

Data Download Date/Time: 07/31/2018 19:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001179

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEREZ,PEREZ,  
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)  
ARRESTING OFFICER: SAMPLE,CARSON  
ARRESTING OFFICER AGENCY: FFWCC (West Palm)  
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/24/2019  
VIOLATION CODE: BUI  
ARREST TIME: 19:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:28

SUBJECT'S LAST NAME: GONZALEZ GONZALEZ

DATE/TIME OF TEST: 07/01/2018 20:03:05

---

Diagnostics Check	OK	19:56
Air Blank	0.000	19:56
Control Test	0.079	19:57
Air Blank	0.000	19:57
Subject Sample # 1	0.112	19:58
Breath Volume	1.496	19:58
Air Blank	0.000	19:59
Air Blank	0.000	20:01
Subject Sample # 2	0.113	20:01
Breath Volume	1.527	20:01
Air Blank	0.000	20:02
Control Test	0.079	20:02
Air Blank	0.000	20:02
Diagnostics Check	OK	20:03

---

Data Download Date/Time: 07/31/2018 19:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001179

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEREZ,ADRIAN,  
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)  
ARRESTING OFFICER: SIMON,CHRISTOPHER  
ARRESTING OFFICER AGENCY: FFWCC (West Palm)  
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/24/2019  
VIOLATION CODE: BUI  
ARREST TIME: 20:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: ENRIQUEZ

DATE/TIME OF TEST: 07/01/2018 20:44:59

---

Diagnostics Check	OK	20:34	
Air Blank	0.000	20:35	
Control Test	0.079	20:35	
Air Blank	0.000	20:36	
Subject Sample	# 1 0.108	20:39	Volume Not Met
Breath Volume	0.000	20:39	
Air Blank	0.000	20:39	
Air Blank	0.000	20:41	
Subject Sample	# 2	20:43	Subject Test Refused
Breath Volume	0.000	20:43	
Air Blank	0.000	20:44	
Control Test	0.079	20:44	
Air Blank	0.000	20:44	
Diagnostics Check	OK	20:44	

---

Data Download Date/Time: 07/31/2018 19:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001179

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PLUSA,KYLE,A  
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)  
ARRESTING OFFICER: STEARNS,MICHAEL  
ARRESTING OFFICER AGENCY: FFWCC (West Palm)  
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/24/2019  
VIOLATION CODE: BUI  
ARREST TIME: 20:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: GOMEZ CAMEJO

DATE/TIME OF TEST: 07/07/2018 20:44:29

---

Diagnostics Check	OK	20:41	
Air Blank	0.000	20:42	
Control Test	0.081	20:42	
Air Blank	0.000	20:42	
Subject Sample	# 1	20:43	Subject Test Refused
Breath Volume	0.000	20:43	
Air Blank	0.000	20:43	
Control Test	0.080	20:43	
Air Blank	0.000	20:44	
Diagnostics Check	OK	20:44	

---

Data Download Date/Time: 07/31/2018 19:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001179

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REYNAUD,GEORGE,  
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)  
ARRESTING OFFICER: BLANCO,JUAN  
ARRESTING OFFICER AGENCY: FFWCC (West Palm)  
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/24/2019  
VIOLATION CODE: BUI  
ARREST TIME: 21:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: MARQUEZ BERMUDEZ

DATE/TIME OF TEST: 07/07/2018 21:28:13

---

Diagnostics Check	OK	21:25	
Air Blank	0.000	21:25	
Control Test	0.076	21:26	
Air Blank	0.000	21:26	
Subject Sample	# 1	21:26	Subject Test Refused
Breath Volume	0.000	21:26	
Air Blank	0.000	21:27	
Control Test	0.076	21:27	
Air Blank	0.000	21:28	
Diagnostics Check	OK	21:28	

---

Data Download Date/Time: 07/31/2018 19:32



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001190

INSTRUMENT REGISTERED TO: BOYNTON BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CASTRO,DENNIS,  
BREATH TEST OPERATOR AGENCY: Boynton Beach PD  
ARRESTING OFFICER: CASTRO,DENNIS  
ARRESTING OFFICER AGENCY: Boynton Beach PD  
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 390314  
DRY GAS EXPIRY DATE: 04/19/2019  
VIOLATION CODE: DUI  
ARREST TIME: 18:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:48

SUBJECT'S LAST NAME: THESSA

DATE/TIME OF TEST: 07/18/2018 19:18:33

---

Diagnostics Check	OK	19:12
Air Blank	0.000	19:12
Control Test	0.080	19:13
Air Blank	0.000	19:13
Subject Sample # 1	0.247	19:14
Breath Volume	2.722	19:14
Air Blank	0.000	19:14
Air Blank	0.000	19:16
Subject Sample # 2	0.251	19:17
Breath Volume	3.097	19:17
Air Blank	0.000	19:17
Control Test	0.080	19:17
Air Blank	0.000	19:18
Diagnostics Check	OK	19:18

---

Data Download Date/Time: 07/27/2018 22:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ARAFA,YUSUF,A  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: ARAFA,YUSUF  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 07/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 572920  
DRY GAS EXPIRY DATE: 09/13/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: BARLOW

DATE/TIME OF TEST: 07/24/2018 02:02:26

---

Diagnostics Check		OK	01:50	
Air Blank		0.000	01:50	
Control Test		0.079	01:51	
Air Blank		0.000	01:51	
Subject Sample	# 1	0.203	01:52	
Breath Volume		1.675	01:52	
Air Blank		0.000	01:52	
Air Blank		0.000	01:54	
Subject Sample	# 2	0.195	01:57	Volume Not Met
Breath Volume		0.566	01:57	
Air Blank		0.000	01:58	
Air Blank		0.000	02:00	
Subject Sample	# 3	0.209	02:00	
Breath Volume		1.742	02:00	
Air Blank		0.000	02:01	
Control Test		0.079	02:01	
Air Blank		0.000	02:02	
Diagnostics Check		OK	02:02	

---

Data Download Date/Time: 08/03/2018 12:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GILLMAN,JAMES,L  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: ORNELAS,INEZ  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 07/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 572920  
DRY GAS EXPIRY DATE: 09/13/2019  
VIOLATION CODE: DUI  
ARREST TIME: 06:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:42

SUBJECT'S LAST NAME: GRUNDY

DATE/TIME OF TEST: 07/30/2018 07:16:11

---

Diagnostics Check	OK	07:09
Air Blank	0.000	07:10
Control Test	0.080	07:10
Air Blank	0.000	07:11
Subject Sample # 1	0.117	07:11
Breath Volume	2.343	07:11
Air Blank	0.000	07:12
Air Blank	0.000	07:14
Subject Sample # 2	0.113	07:14
Breath Volume	1.937	07:14
Air Blank	0.000	07:15
Control Test	0.079	07:15
Air Blank	0.000	07:16
Diagnostics Check	OK	07:16

---

Data Download Date/Time: 08/03/2018 12:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ABRAHAM,ERIC,D  
BREATH TEST OPERATOR AGENCY: Panama City PD  
ARRESTING OFFICER: ABRAHAM,ERIC  
ARRESTING OFFICER AGENCY: Panama City PD  
LAST AGENCY INSPECTION: 07/15/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 366849  
DRY GAS EXPIRY DATE: 03/30/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:18

SUBJECT'S LAST NAME: HAMILTON

DATE/TIME OF TEST: 07/15/2018 21:58:35

---

Diagnostics Check		OK	21:52
Air Blank		0.000	21:52
Control Test		0.080	21:52
Air Blank		0.000	21:53
Subject Sample	# 1	0.226	21:53
Breath Volume		1.960	21:53
Air Blank		0.000	21:54
Air Blank		0.000	21:56
Subject Sample	# 2	0.224	21:57
Breath Volume		2.796	21:57
Air Blank		0.000	21:57
Control Test		0.079	21:58
Air Blank		0.000	21:58
Diagnostics Check		OK	21:58

---

Data Download Date/Time: 08/28/2018 21:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MERRITT, DAKOTA, A  
BREATH TEST OPERATOR AGENCY: Panama City PD  
ARRESTING OFFICER: SMITH, VICTORIA  
ARRESTING OFFICER AGENCY: Panama City PD  
LAST AGENCY INSPECTION: 07/15/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 366849  
DRY GAS EXPIRY DATE: 03/30/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 07/17/2018 22:17:08

---

Diagnostics Check	OK	22:10
Air Blank	0.000	22:11
Control Test	0.080	22:11
Air Blank	0.000	22:12
Subject Sample # 1	0.000	22:12
Breath Volume	2.621	22:12
Air Blank	0.000	22:13
Air Blank	0.000	22:15
Subject Sample # 2	0.000	22:15
Breath Volume	2.636	22:15
Air Blank	0.000	22:16
Control Test	0.080	22:16
Air Blank	0.000	22:17
Diagnostics Check	OK	22:17

---

Data Download Date/Time: 08/28/2018 21:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TURNER,JONATHAN,R  
BREATH TEST OPERATOR AGENCY: Panama City PD  
ARRESTING OFFICER: HEALD,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Panama City PD  
LAST AGENCY INSPECTION: 07/15/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 366849  
DRY GAS EXPIRY DATE: 03/30/2019  
VIOLATION CODE: DUI  
ARREST TIME: 04:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:39

SUBJECT'S LAST NAME: GIBSON

DATE/TIME OF TEST: 07/20/2018 05:05:34

---

Diagnostics Check	OK	04:59
Air Blank	0.000	04:59
Control Test	0.081	05:00
Air Blank	0.000	05:00
Subject Sample # 1	0.066	05:01
Breath Volume	1.250	05:01
Air Blank	0.000	05:01
Air Blank	0.000	05:03
Subject Sample # 2	0.067	05:04
Breath Volume	1.640	05:04
Air Blank	0.000	05:04
Control Test	0.080	05:05
Air Blank	0.000	05:05
Diagnostics Check	OK	05:05

---

Data Download Date/Time: 08/28/2018 21:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ABRAHAM,ERIC,D  
BREATH TEST OPERATOR AGENCY: Panama City PD  
ARRESTING OFFICER: ABRAHAM,ERIC  
ARRESTING OFFICER AGENCY: Panama City PD  
LAST AGENCY INSPECTION: 07/15/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 366849  
DRY GAS EXPIRY DATE: 03/30/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:13  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:13

SUBJECT'S LAST NAME: DODGEN-EVANS

DATE/TIME OF TEST: 07/21/2018 22:19:47

---

Diagnostics Check	OK	22:15	
Air Blank	0.000	22:15	
Control Test	0.080	22:16	
Air Blank	0.000	22:16	
Subject Sample	# 1	22:18	Subject Test Refused
Breath Volume	0.000	22:18	
Air Blank	0.000	22:18	
Control Test	0.081	22:19	
Air Blank	0.000	22:19	
Diagnostics Check	OK	22:19	

---

Data Download Date/Time: 08/28/2018 21:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: QUINN,CAMERON,B  
BREATH TEST OPERATOR AGENCY: Panama City PD  
ARRESTING OFFICER: QUINN,CAMERON  
ARRESTING OFFICER AGENCY: Panama City PD  
LAST AGENCY INSPECTION: 07/15/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 366849  
DRY GAS EXPIRY DATE: 03/30/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 07/29/2018 19:34:13

---

Diagnostics Check	OK	19:32	
Air Blank	0.000	19:32	
Control Test	0.080	19:33	
Air Blank	0.000	19:33	
Subject Sample	# 1	19:33	Sequence Aborted
Breath Volume	0.000	19:33	
Air Blank	0.000	19:34	

---

Data Download Date/Time: 08/28/2018 21:21



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOHRER,MARK,  
BREATH TEST OPERATOR AGENCY: Panama City PD  
ARRESTING OFFICER: BOHRER,MARK  
ARRESTING OFFICER AGENCY: Panama City PD  
LAST AGENCY INSPECTION: 07/15/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 366849  
DRY GAS EXPIRY DATE: 03/30/2019  
VIOLATION CODE: DUI  
ARREST TIME: 19:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: STRYCHALSKI

DATE/TIME OF TEST: 07/29/2018 19:42:16

---

Diagnostics Check	OK	19:36
Air Blank	0.000	19:36
Control Test	0.080	19:36
Air Blank	0.000	19:37
Subject Sample # 1	0.330	19:37
Breath Volume	2.152	19:37
Air Blank	0.000	19:38
Air Blank	0.000	19:40
Subject Sample # 2	0.327	19:40
Breath Volume	1.578	19:40
Air Blank	0.000	19:41
Control Test	0.080	19:41
Air Blank	0.000	19:42
Diagnostics Check	OK	19:42

---

Data Download Date/Time: 08/28/2018 21:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001207

INSTRUMENT REGISTERED TO: HENDRY COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NORWOOD,CHRISTOPHER,J  
BREATH TEST OPERATOR AGENCY: Hendry County SO  
ARRESTING OFFICER: NORWOOD,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Hendry County SO  
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 83551  
DRY GAS EXPIRY DATE: 10/02/2018  
VIOLATION CODE: DUI  
ARREST TIME: 19:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:42

SUBJECT'S LAST NAME: SAMLALL

DATE/TIME OF TEST: 07/26/2018 20:08:55

---

Diagnostics Check		OK	20:02
Air Blank		0.000	20:02
Control Test		0.081	20:03
Air Blank		0.000	20:03
Subject Sample	# 1	0.246	20:04
Breath Volume		1.996	20:04
Air Blank		0.000	20:04
Air Blank		0.000	20:06
Subject Sample	# 2	0.240	20:07
Breath Volume		1.976	20:07
Air Blank		0.000	20:07
Control Test		0.081	20:08
Air Blank		0.000	20:08
Diagnostics Check		OK	20:08

---

Data Download Date/Time: 07/27/2018 22:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001207

INSTRUMENT REGISTERED TO: HENDRY COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NORWOOD,CHRISTOPHER,J  
BREATH TEST OPERATOR AGENCY: Hendry County SO  
ARRESTING OFFICER: CORONILLA,DANIEL  
ARRESTING OFFICER AGENCY: Hendry County SO  
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 83551  
DRY GAS EXPIRY DATE: 10/02/2018  
VIOLATION CODE: DUI  
ARREST TIME: 22:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: MARTINEZ CARCASSES

DATE/TIME OF TEST: 07/26/2018 22:47:40

---

Diagnostics Check	OK	22:40
Air Blank	0.000	22:41
Control Test	0.081	22:41
Air Blank	0.000	22:41
Subject Sample # 1	0.177	22:42
Breath Volume	1.527	22:42
Air Blank	0.000	22:42
Air Blank	0.000	22:44
Subject Sample # 2	0.171	22:45
Breath Volume	1.468	22:45
Air Blank	0.000	22:46
Control Test	0.083	22:47
Air Blank	0.000	22:47
Diagnostics Check	OK	22:47

---

Data Download Date/Time: 07/27/2018 22:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001207

INSTRUMENT REGISTERED TO: HENDRY COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NORWOOD,CHRISTOPHER,J  
BREATH TEST OPERATOR AGENCY: Hendry County SO  
ARRESTING OFFICER: SHAW,ROBERT  
ARRESTING OFFICER AGENCY: Clewiston PD  
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 83551  
DRY GAS EXPIRY DATE: 10/02/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 07/27/2018 01:45:01

---

Diagnostics Check	OK	01:42	
Air Blank	0.000	01:42	
Control Test	0.081	01:42	
Air Blank	0.000	01:43	
Subject Sample	# 1	01:43	Subject Test Refused
Breath Volume	0.000	01:43	
Air Blank	0.000	01:44	
Control Test	0.081	01:44	
Air Blank	0.000	01:44	
Diagnostics Check	OK	01:45	

---

Data Download Date/Time: 07/27/2018 22:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

INSTRUMENT REGISTERED TO: TYNDALL AFB  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SOUZA,AARON,B  
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD  
ARRESTING OFFICER: SASTRE,JAMES  
ARRESTING OFFICER AGENCY: Tyndall AFB PD  
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32316080A3  
DRY GAS EXPIRY DATE: 01/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 13:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:20

SUBJECT'S LAST NAME: BUSH

DATE/TIME OF TEST: 07/24/2018 13:52:51

---

Diagnostics Check	OK	13:44
Air Blank	0.000	13:44
Control Test	0.077	13:45
Air Blank	0.000	13:45
Subject Sample # 1	0.247	13:46
Breath Volume	1.789	13:46
Air Blank	0.000	13:46
Air Blank	0.000	13:48
Subject Sample # 2	0.230	13:51
Breath Volume	1.984	13:51
Air Blank	0.000	13:51
Control Test	0.078	13:52
Air Blank	0.000	13:52
Diagnostics Check	OK	13:52

---

Data Download Date/Time: 08/01/2018 10:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

INSTRUMENT REGISTERED TO: AUBURNDALE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BAKER,MATT,A  
BREATH TEST OPERATOR AGENCY: Auburndale PD  
ARRESTING OFFICER: HAVRANEK,ADAM  
ARRESTING OFFICER AGENCY: Auburndale PD  
LAST AGENCY INSPECTION: 06/26/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03417080A3  
DRY GAS EXPIRY DATE: 03/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: WARD

DATE/TIME OF TEST: 07/04/2018 03:50:55

---

Diagnostics Check	OK	03:47	
Air Blank	0.000	03:48	
Control Test	0.080	03:48	
Air Blank	0.000	03:49	
Subject Sample	# 1	03:49	Subject Test Refused
Breath Volume	0.000	03:49	
Air Blank	0.000	03:50	
Control Test	0.080	03:50	
Air Blank	0.000	03:50	
Diagnostics Check	OK	03:50	

---

Data Download Date/Time: 07/26/2018 11:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

INSTRUMENT REGISTERED TO: AUBURNDALE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BAKER, MATTHEW, A  
BREATH TEST OPERATOR AGENCY: Auburndale PD  
ARRESTING OFFICER: PENA, A  
ARRESTING OFFICER AGENCY: Auburndale PD  
LAST AGENCY INSPECTION: 06/26/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03417080A3  
DRY GAS EXPIRY DATE: 03/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 07/21/2018 01:36:22

---

Diagnostics Check		OK	01:29
Air Blank		0.000	01:30
Control Test		0.079	01:30
Air Blank		0.000	01:31
Subject Sample	# 1	0.230	01:31
Breath Volume		1.375	01:31
Air Blank		0.000	01:32
Air Blank		0.000	01:34
Subject Sample	# 2	0.245	01:34
Breath Volume		2.617	01:34
Air Blank		0.000	01:35
Control Test		0.078	01:35
Air Blank		0.000	01:36
Diagnostics Check		OK	01:36

---

Data Download Date/Time: 07/26/2018 11:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SECORD,RYAN,T  
BREATH TEST OPERATOR AGENCY: West Palm Beach PD  
ARRESTING OFFICER: DILLARD,DANIEL  
ARRESTING OFFICER AGENCY: West Palm Beach PD  
LAST AGENCY INSPECTION: 06/23/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 152169  
DRY GAS EXPIRY DATE: 10/30/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:53  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: HURSKY

DATE/TIME OF TEST: 07/01/2018 04:05:31

---

Diagnostics Check	OK	04:02	
Air Blank	0.000	04:03	
Control Test	0.080	04:03	
Air Blank	0.000	04:03	
Subject Sample	# 1	04:04	Subject Test Refused
Breath Volume	0.000	04:04	
Air Blank	0.000	04:04	
Control Test	0.080	04:04	
Air Blank	0.000	04:05	
Diagnostics Check	OK	04:05	

---

Data Download Date/Time: 07/30/2018 12:51



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POTTER,ADAM,H  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: POTTER,ADAM  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 17:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: PEREZ MALDONADO

DATE/TIME OF TEST: 07/04/2018 18:33:01

---

Diagnostics Check	OK	18:26
Air Blank	0.000	18:26
Control Test	0.080	18:27
Air Blank	0.000	18:27
Subject Sample # 1	0.131	18:28
Breath Volume	2.996	18:28
Air Blank	0.000	18:29
Air Blank	0.000	18:30
Subject Sample # 2	0.126	18:31
Breath Volume	2.398	18:31
Air Blank	0.000	18:32
Control Test	0.079	18:32
Air Blank	0.000	18:32
Diagnostics Check	OK	18:33

---

Data Download Date/Time: 07/27/2018 10:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POTTER,ADAM,H  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: POTTER,ADAM  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 20:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:41

SUBJECT'S LAST NAME: CREWS

DATE/TIME OF TEST: 07/14/2018 21:08:15

---

Diagnostics Check	OK	21:01
Air Blank	0.000	21:02
Control Test	0.080	21:02
Air Blank	0.000	21:03
Subject Sample # 1	0.108	21:03
Breath Volume	2.843	21:03
Air Blank	0.000	21:04
Air Blank	0.000	21:06
Subject Sample # 2	0.105	21:06
Breath Volume	2.750	21:06
Air Blank	0.000	21:07
Control Test	0.079	21:07
Air Blank	0.000	21:08
Diagnostics Check	OK	21:08

---

Data Download Date/Time: 07/27/2018 10:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SYKES,STEVEN,  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: SYKES,STEVEN  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: VREELAND

DATE/TIME OF TEST: 07/18/2018 04:11:41

---

Diagnostics Check	OK	04:00	
Air Blank	0.000	04:00	
Control Test	0.081	04:01	
Air Blank	0.000	04:01	
Subject Sample	# 1	04:04	No Sample Provided
Breath Volume	0.000	04:04	
Air Blank	0.000	04:05	
Air Blank	0.000	04:07	
Subject Sample	# 2	04:10	No Sample Provided
Breath Volume	0.000	04:10	
Air Blank	0.000	04:10	
Control Test	0.080	04:11	
Air Blank	0.000	04:11	
Diagnostics Check	OK	04:11	

---

Data Download Date/Time: 07/27/2018 10:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FARRELL,JEREMY,D  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: RESENDEZ,GILVRANDO  
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)  
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: BUI  
ARREST TIME: 15:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:10

SUBJECT'S LAST NAME: THAYER

DATE/TIME OF TEST: 07/21/2018 16:42:49

---

Diagnostics Check	OK	16:36
Air Blank	0.000	16:37
Control Test	0.079	16:37
Air Blank	0.000	16:38
Subject Sample # 1	0.120	16:38
Breath Volume	1.414	16:38
Air Blank	0.000	16:39
Air Blank	0.000	16:40
Subject Sample # 2	0.127	16:41
Breath Volume	1.421	16:41
Air Blank	0.000	16:41
Control Test	0.079	16:42
Air Blank	0.000	16:42
Diagnostics Check	OK	16:42

---

Data Download Date/Time: 07/27/2018 10:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WARREN,BILLY,J  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: WARREN,BILLY  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: GAMEZ

DATE/TIME OF TEST: 07/22/2018 23:59:23

---

Diagnostics Check		OK	23:47
Air Blank		0.000	23:47
Control Test		0.080	23:48
Air Blank		0.000	23:48
Subject Sample	# 1	0.211	23:50
Breath Volume		2.230	23:50
Air Blank		0.000	23:51
Air Blank		0.000	23:53
Subject Sample	# 2	0.189	23:54
Breath Volume		1.792	23:54
Air Blank		0.000	23:54
Air Blank		0.000	23:56
Subject Sample	# 3	0.192	23:57
Breath Volume		2.000	23:57
Air Blank		0.000	23:58
Control Test		0.079	23:58
Air Blank		0.000	23:59
Diagnostics Check		OK	23:59

---

Data Download Date/Time: 07/27/2018 10:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LAWES,ERIC,  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: LAWES,ERIC  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 07/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:41  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: MOLINA ORSINI

DATE/TIME OF TEST: 07/28/2018 04:27:59

---

Diagnostics Check	OK	04:21
Air Blank	0.000	04:22
Control Test	0.079	04:22
Air Blank	0.000	04:22
Subject Sample # 1	0.032	04:23
Breath Volume	2.503	04:23
Air Blank	0.000	04:24
Air Blank	0.000	04:25
Subject Sample # 2	0.032	04:26
Breath Volume	2.542	04:26
Air Blank	0.000	04:27
Control Test	0.079	04:27
Air Blank	0.000	04:27
Diagnostics Check	OK	04:27

---

Data Download Date/Time: 08/29/2018 13:57

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LAWES,ERIC,  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: ROBERTS,DALE  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 07/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: MONTEPEQUE

DATE/TIME OF TEST: 07/28/2018 04:46:44

---

Diagnostics Check		OK	04:40
Air Blank		0.000	04:40
Control Test		0.079	04:41
Air Blank		0.000	04:41
Subject Sample	# 1	0.140	04:42
Breath Volume		2.074	04:42
Air Blank		0.000	04:42
Air Blank		0.000	04:44
Subject Sample	# 2	0.138	04:45
Breath Volume		2.351	04:45
Air Blank		0.000	04:45
Control Test		0.079	04:46
Air Blank		0.000	04:46
Diagnostics Check		OK	04:46

---

Data Download Date/Time: 08/29/2018 13:57

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MELTON,XAVIER,V  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: ROBERTS,DALE  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 07/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 17:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:04

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 07/28/2018 18:39:16

---

Diagnostics Check		OK	18:32
Air Blank		0.000	18:33
Control Test		0.080	18:33
Air Blank		0.000	18:34
Subject Sample	# 1	0.273	18:34
Breath Volume		2.707	18:34
Air Blank		0.000	18:35
Air Blank		0.000	18:37
Subject Sample	# 2	0.272	18:37
Breath Volume		3.292	18:37
Air Blank		0.000	18:38
Control Test		0.078	18:38
Air Blank		0.000	18:39
Diagnostics Check		OK	18:39

---

Data Download Date/Time: 08/29/2018 13:57



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MELTON,XAVIER,V  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: MCKEE,JOHN  
ARRESTING OFFICER AGENCY: Bartow PD  
LAST AGENCY INSPECTION: 07/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:26  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: PARKS

DATE/TIME OF TEST: 07/28/2018 23:29:28

---

Diagnostics Check	OK	23:23
Air Blank	0.000	23:23
Control Test	0.079	23:23
Air Blank	0.000	23:24
Subject Sample # 1	0.199	23:25
Breath Volume	1.914	23:25
Air Blank	0.000	23:25
Air Blank	0.000	23:27
Subject Sample # 2	0.203	23:27
Breath Volume	1.949	23:27
Air Blank	0.000	23:28
Control Test	0.078	23:28
Air Blank	0.000	23:29
Diagnostics Check	OK	23:29

---

Data Download Date/Time: 08/29/2018 13:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MELTON,XAVIER,V  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: ROBERTS,DALE  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 07/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: ROSALES

DATE/TIME OF TEST: 07/29/2018 02:11:35

---

Diagnostics Check		OK	02:05
Air Blank		0.000	02:05
Control Test		0.079	02:06
Air Blank		0.000	02:06
Subject Sample	# 1	0.288	02:07
Breath Volume		2.988	02:07
Air Blank		0.000	02:07
Air Blank		0.000	02:09
Subject Sample	# 2	0.277	02:09
Breath Volume		1.164	02:09
Air Blank		0.000	02:10
Control Test		0.077	02:10
Air Blank		0.000	02:11
Diagnostics Check		OK	02:11

---

Data Download Date/Time: 08/29/2018 13:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MELTON,XAVIER,V  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: ROBERTS,DALE  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 07/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: DEEMER

DATE/TIME OF TEST: 07/30/2018 01:20:24

---

Diagnostics Check	OK	01:17	
Air Blank	0.000	01:18	
Control Test	0.079	01:18	
Air Blank	0.000	01:18	
Subject Sample	# 1	01:19	Subject Test Refused
Breath Volume	0.000	01:19	
Air Blank	0.000	01:19	
Control Test	0.080	01:19	
Air Blank	0.000	01:20	
Diagnostics Check	OK	01:20	

---

Data Download Date/Time: 08/29/2018 13:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PITCHER,DARRYL,P  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: BINDER,BRONSON  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814181  
DRY GAS EXPIRY DATE: 07/24/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: HERRIN

DATE/TIME OF TEST: 07/05/2018 00:22:59

---

Diagnostics Check	OK	00:14
Air Blank	0.000	00:15
Control Test	0.080	00:15
Air Blank	0.000	00:16
Subject Sample # 1	0.112	00:18
Breath Volume	2.761	00:18
Air Blank	0.000	00:19
Air Blank	0.000	00:20
Subject Sample # 2	0.106	00:21
Breath Volume	2.433	00:21
Air Blank	0.000	00:22
Control Test	0.079	00:22
Air Blank	0.000	00:22
Diagnostics Check	OK	00:22

---

Data Download Date/Time: 07/06/2018 23:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PITCHER,DARRYL,P  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: BINDER,BRONSON  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814181  
DRY GAS EXPIRY DATE: 07/24/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:33  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:31

SUBJECT'S LAST NAME: EMMETT

DATE/TIME OF TEST: 07/05/2018 20:58:44

---

Diagnostics Check	OK	20:52
Air Blank	0.000	20:52
Control Test	0.080	20:52
Air Blank	0.000	20:53
Subject Sample # 1	0.037	20:54
Breath Volume	2.613	20:54
Air Blank	0.000	20:54
Air Blank	0.000	20:56
Subject Sample # 2	0.035	20:57
Breath Volume	1.972	20:57
Air Blank	0.000	20:57
Control Test	0.081	20:58
Air Blank	0.000	20:58
Diagnostics Check	OK	20:58

---

Data Download Date/Time: 07/06/2018 23:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PITCHER,DARRYL,P  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: BINDER,BRONSON  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814181  
DRY GAS EXPIRY DATE: 07/24/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: SAVAGE

DATE/TIME OF TEST: 07/06/2018 02:37:54

---

Diagnostics Check	OK	02:30
Air Blank	0.000	02:31
Control Test	0.082	02:31
Air Blank	0.000	02:32
Subject Sample # 1	0.095	02:33
Breath Volume	3.000	02:33
Air Blank	0.000	02:33
Air Blank	0.000	02:35
Subject Sample # 2	0.091	02:36
Breath Volume	2.003	02:36
Air Blank	0.000	02:37
Control Test	0.078	02:37
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

---

Data Download Date/Time: 07/06/2018 23:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALBRIDGE,CODY,A  
BREATH TEST OPERATOR AGENCY: Umatilla PD  
ARRESTING OFFICER: ALBRIDGE,CODY  
ARRESTING OFFICER AGENCY: Umatilla PD  
LAST AGENCY INSPECTION: 07/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814181  
DRY GAS EXPIRY DATE: 07/24/2020  
VIOLATION CODE: DUI  
ARREST TIME: 04:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: MYER

DATE/TIME OF TEST: 07/07/2018 05:19:14

---

Diagnostics Check	OK	05:12
Air Blank	0.000	05:12
Control Test	0.081	05:13
Air Blank	0.000	05:13
Subject Sample # 1	0.147	05:14
Breath Volume	1.199	05:14
Air Blank	0.000	05:14
Air Blank	0.000	05:16
Subject Sample # 2	0.153	05:17
Breath Volume	3.265	05:17
Air Blank	0.000	05:18
Control Test	0.080	05:18
Air Blank	0.000	05:19
Diagnostics Check	OK	05:19

---

Data Download Date/Time: 08/15/2018 23:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOTZ,KATHRYN,N  
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores  
ARRESTING OFFICER: GOTZ,KATHRYN  
ARRESTING OFFICER AGENCY: Daytona Beach Shores  
LAST AGENCY INSPECTION: 06/20/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 906364  
DRY GAS EXPIRY DATE: 02/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:48

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 07/10/2018 21:16:47

---

Diagnostics Check	OK	21:10
Air Blank	0.000	21:11
Control Test	0.079	21:11
Air Blank	0.000	21:12
Subject Sample # 1	0.000	21:12
Breath Volume	1.914	21:12
Air Blank	0.000	21:13
Air Blank	0.000	21:15
Subject Sample # 2	0.000	21:15
Breath Volume	1.761	21:15
Air Blank	0.000	21:15
Control Test	0.080	21:16
Air Blank	0.000	21:16
Diagnostics Check	OK	21:16

---

Data Download Date/Time: 08/03/2018 20:51



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CASTELLANO,VINCENT,J  
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores  
ARRESTING OFFICER: CASTELLANO,VINCENT  
ARRESTING OFFICER AGENCY: Daytona Beach Shores  
LAST AGENCY INSPECTION: 06/20/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 906364  
DRY GAS EXPIRY DATE: 02/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:12  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: MOODY

DATE/TIME OF TEST: 07/15/2018 01:06:16

---

Diagnostics Check	OK	00:59
Air Blank	0.000	01:00
Control Test	0.080	01:00
Air Blank	0.000	01:01
Subject Sample # 1	0.061	01:01
Breath Volume	2.328	01:01
Air Blank	0.000	01:02
Air Blank	0.000	01:04
Subject Sample # 2	0.059	01:04
Breath Volume	2.199	01:04
Air Blank	0.000	01:05
Control Test	0.079	01:05
Air Blank	0.000	01:06
Diagnostics Check	OK	01:06

---

Data Download Date/Time: 08/03/2018 20:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOTZ,KATHRYN,N  
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores  
ARRESTING OFFICER: GOTZ,KATHRYN  
ARRESTING OFFICER AGENCY: Daytona Beach Shores  
LAST AGENCY INSPECTION: 06/20/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 906364  
DRY GAS EXPIRY DATE: 02/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: JUDY

DATE/TIME OF TEST: 07/29/2018 01:29:41

---

Diagnostics Check	OK	01:22
Air Blank	0.000	01:23
Control Test	0.080	01:23
Air Blank	0.000	01:23
Subject Sample # 1	0.031	01:24
Breath Volume	1.148	01:24
Air Blank	0.000	01:25
Air Blank	0.000	01:27
Subject Sample # 2	0.036	01:28
Breath Volume	2.285	01:28
Air Blank	0.000	01:28
Control Test	0.080	01:29
Air Blank	0.000	01:29
Diagnostics Check	OK	01:29

---

Data Download Date/Time: 08/03/2018 20:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BINZ, ANTHONY, E  
BREATH TEST OPERATOR AGENCY: Edgewater PD  
ARRESTING OFFICER: BUEHLER, ANTHONY  
ARRESTING OFFICER AGENCY: Edgewater PD  
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 765462  
DRY GAS EXPIRY DATE: 04/13/2020  
VIOLATION CODE: DUI  
ARREST TIME: 18:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:19

SUBJECT'S LAST NAME: HASTINGS

DATE/TIME OF TEST: 07/02/2018 18:42:13

---

Diagnostics Check	OK	18:39	
Air Blank	0.000	18:39	
Control Test	0.078	18:39	
Air Blank	0.000	18:40	
Subject Sample	# 1	18:40	Subject Test Refused
Breath Volume	0.000	18:40	
Air Blank	0.000	18:41	
Control Test	0.077	18:41	
Air Blank	0.000	18:42	
Diagnostics Check	OK	18:42	

---

Data Download Date/Time: 08/18/2018 23:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FIELDS,ERIC,R  
BREATH TEST OPERATOR AGENCY: Port Orange PD  
ARRESTING OFFICER: SHLEIFER,BRIAN  
ARRESTING OFFICER AGENCY: Port Orange PD  
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 609878  
DRY GAS EXPIRY DATE: 11/01/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:28  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: GILCHRIST

DATE/TIME OF TEST: 07/12/2018 22:15:28

---

Diagnostics Check	OK	22:09
Air Blank	0.000	22:09
Control Test	0.080	22:09
Air Blank	0.000	22:10
Subject Sample # 1	0.146	22:10
Breath Volume	1.824	22:10
Air Blank	0.000	22:11
Air Blank	0.000	22:13
Subject Sample # 2	0.134	22:13
Breath Volume	1.804	22:13
Air Blank	0.000	22:14
Control Test	0.081	22:14
Air Blank	0.000	22:15
Diagnostics Check	OK	22:15

---

Data Download Date/Time: 08/11/2018 01:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FIELDS,ERIC,R  
BREATH TEST OPERATOR AGENCY: Port Orange PD  
ARRESTING OFFICER: ALEVISATOS,JARRAD  
ARRESTING OFFICER AGENCY: Port Orange PD  
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 609878  
DRY GAS EXPIRY DATE: 11/01/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:07  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: ELLIOT

DATE/TIME OF TEST: 07/22/2018 01:56:45

---

Diagnostics Check	OK	01:50
Air Blank	0.000	01:50
Control Test	0.081	01:50
Air Blank	0.000	01:51
Subject Sample # 1	0.142	01:52
Breath Volume	1.824	01:52
Air Blank	0.000	01:52
Air Blank	0.000	01:54
Subject Sample # 2	0.132	01:55
Breath Volume	1.402	01:55
Air Blank	0.000	01:55
Control Test	0.080	01:56
Air Blank	0.000	01:56
Diagnostics Check	OK	01:56

---

Data Download Date/Time: 08/11/2018 01:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KURTZ,THEODORE,F  
BREATH TEST OPERATOR AGENCY: Port Orange PD  
ARRESTING OFFICER: SHLEIFER,BRIAN  
ARRESTING OFFICER AGENCY: Port Orange PD  
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 609878  
DRY GAS EXPIRY DATE: 11/01/2019  
VIOLATION CODE: DUI  
ARREST TIME: 04:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: BRACEWELL

DATE/TIME OF TEST: 07/22/2018 05:14:44

---

Diagnostics Check	OK	05:08
Air Blank	0.000	05:08
Control Test	0.082	05:09
Air Blank	0.000	05:09
Subject Sample # 1	0.077	05:10
Breath Volume	2.160	05:10
Air Blank	0.000	05:10
Air Blank	0.000	05:12
Subject Sample # 2	0.078	05:13
Breath Volume	1.980	05:13
Air Blank	0.000	05:13
Control Test	0.082	05:14
Air Blank	0.000	05:14
Diagnostics Check	OK	05:14

---

Data Download Date/Time: 08/11/2018 01:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MANDESE,ANGELO,  
BREATH TEST OPERATOR AGENCY: Port Orange PD  
ARRESTING OFFICER: SLUDER,JAMES  
ARRESTING OFFICER AGENCY: Port Orange PD  
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 609878  
DRY GAS EXPIRY DATE: 11/01/2019  
VIOLATION CODE: DUI  
ARREST TIME: 20:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: JENSON

DATE/TIME OF TEST: 07/24/2018 21:53:00

---

Diagnostics Check	OK	21:46
Air Blank	0.000	21:47
Control Test	0.082	21:47
Air Blank	0.000	21:47
Subject Sample # 1	0.229	21:48
Breath Volume	2.394	21:48
Air Blank	0.000	21:49
Air Blank	0.000	21:50
Subject Sample # 2	0.221	21:51
Breath Volume	1.777	21:51
Air Blank	0.000	21:52
Control Test	0.081	21:52
Air Blank	0.000	21:52
Diagnostics Check	OK	21:53

---

Data Download Date/Time: 08/11/2018 01:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAMARA,CRYSTA,N  
BREATH TEST OPERATOR AGENCY: Port Orange PD  
ARRESTING OFFICER: CAMARA,CRYSTA  
ARRESTING OFFICER AGENCY: Port Orange PD  
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 609878  
DRY GAS EXPIRY DATE: 11/01/2019  
VIOLATION CODE: DUI  
ARREST TIME: 17:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: JOHNSTON JR

DATE/TIME OF TEST: 07/26/2018 18:27:40

---

Diagnostics Check	OK	18:21
Air Blank	0.000	18:21
Control Test	0.081	18:22
Air Blank	0.000	18:22
Subject Sample # 1	0.000	18:23
Breath Volume	4.695	18:23
Air Blank	0.000	18:23
Air Blank	0.000	18:25
Subject Sample # 2	0.000	18:26
Breath Volume	4.789	18:26
Air Blank	0.000	18:26
Control Test	0.081	18:27
Air Blank	0.000	18:27
Diagnostics Check	OK	18:27

---

Data Download Date/Time: 08/11/2018 01:30



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FIELDS,ERIC,R  
BREATH TEST OPERATOR AGENCY: Port Orange PD  
ARRESTING OFFICER: SHLEIFER,BRIAN  
ARRESTING OFFICER AGENCY: Port Orange PD  
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 609878  
DRY GAS EXPIRY DATE: 11/01/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:23

SUBJECT'S LAST NAME: FREENY

DATE/TIME OF TEST: 07/26/2018 22:55:29

---

Diagnostics Check	OK	22:49
Air Blank	0.000	22:49
Control Test	0.080	22:50
Air Blank	0.000	22:50
Subject Sample # 1	0.129	22:51
Breath Volume	2.523	22:51
Air Blank	0.000	22:51
Air Blank	0.000	22:53
Subject Sample # 2	0.120	22:54
Breath Volume	1.589	22:54
Air Blank	0.000	22:54
Control Test	0.080	22:54
Air Blank	0.000	22:55
Diagnostics Check	OK	22:55

---

Data Download Date/Time: 08/11/2018 01:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EDWARDS,JUSTIN,D  
BREATH TEST OPERATOR AGENCY: Port Orange PD  
ARRESTING OFFICER: TURCHON,ANTHONY  
ARRESTING OFFICER AGENCY: Port Orange PD  
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 609878  
DRY GAS EXPIRY DATE: 11/01/2019  
VIOLATION CODE: DUI  
ARREST TIME: 16:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:24

SUBJECT'S LAST NAME: ARMER

DATE/TIME OF TEST: 07/30/2018 17:58:22

---

Diagnostics Check		OK	17:48
Air Blank		0.000	17:48
Control Test		0.081	17:48
Air Blank		0.000	17:49
Subject Sample	# 1	0.346	17:49
Breath Volume		1.320	17:49
Air Blank		0.000	17:50
Air Blank		0.000	17:52
Subject Sample	# 2	0.379	17:53
Breath Volume		3.261	17:53
Air Blank		0.000	17:54
Air Blank		0.000	17:56
Subject Sample	# 3	0.373	17:56
Breath Volume		3.121	17:56
Air Blank		0.000	17:57
Control Test		0.078	17:57
Air Blank		0.000	17:58
Diagnostics Check		OK	17:58

---

Data Download Date/Time: 08/11/2018 01:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KURTZ,THEODORE,F  
BREATH TEST OPERATOR AGENCY: Port Orange PD  
ARRESTING OFFICER: SHLEIFER,BRIAN  
ARRESTING OFFICER AGENCY: Port Orange PD  
LAST AGENCY INSPECTION: 07/31/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 609878  
DRY GAS EXPIRY DATE: 11/01/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 07/31/2018 23:57:05

---

Diagnostics Check	OK	23:50
Air Blank	0.000	23:51
Control Test	0.081	23:51
Air Blank	0.000	23:51
Subject Sample # 1	0.000	23:52
Breath Volume	2.097	23:52
Air Blank	0.000	23:53
Air Blank	0.000	23:55
Subject Sample # 2	0.000	23:55
Breath Volume	1.585	23:55
Air Blank	0.000	23:56
Control Test	0.081	23:56
Air Blank	0.000	23:57
Diagnostics Check	OK	23:57

---

Data Download Date/Time: 08/11/2018 01:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HANNON,LAUREN,F  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: ROBIN,GASTON  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 06/13/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19817080A3  
DRY GAS EXPIRY DATE: 09/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: SAUNDERS

DATE/TIME OF TEST: 07/05/2018 02:39:10

---

Diagnostics Check		OK	02:32
Air Blank		0.000	02:33
Control Test		0.080	02:33
Air Blank		0.000	02:33
Subject Sample	# 1	0.200	02:34
Breath Volume		1.359	02:34
Air Blank		0.000	02:34
Air Blank		0.000	02:36
Subject Sample	# 2	0.210	02:37
Breath Volume		3.078	02:37
Air Blank		0.000	02:38
Control Test		0.080	02:38
Air Blank		0.000	02:39
Diagnostics Check		OK	02:39

---

Data Download Date/Time: 07/26/2018 19:52

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARRITY, THOMAS, A  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: GARRITY, THOMAS  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 06/13/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19817080A3  
DRY GAS EXPIRY DATE: 09/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:24  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: CHRISTMAS

DATE/TIME OF TEST: 07/07/2018 23:17:37

---

Diagnostics Check		OK	23:11
Air Blank		0.000	23:11
Control Test		0.081	23:12
Air Blank		0.000	23:12
Subject Sample	# 1	0.242	23:13
Breath Volume		2.660	23:13
Air Blank		0.000	23:13
Air Blank		0.000	23:15
Subject Sample	# 2	0.237	23:16
Breath Volume		2.203	23:16
Air Blank		0.000	23:16
Control Test		0.080	23:17
Air Blank		0.000	23:17
Diagnostics Check		OK	23:17

---

Data Download Date/Time: 07/26/2018 19:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARRITY, THOMAS, A  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: GARRITY, THOMAS  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 06/13/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19817080A3  
DRY GAS EXPIRY DATE: 09/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: DURHAM

DATE/TIME OF TEST: 07/08/2018 02:45:53

---

Diagnostics Check	OK	02:36
Air Blank	0.000	02:37
Control Test	0.080	02:37
Air Blank	0.000	02:37
Subject Sample # 1	0.145	02:41
Breath Volume	1.160	02:41
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.139	02:44
Breath Volume	1.476	02:44
Air Blank	0.000	02:44
Control Test	0.079	02:45
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

---

Data Download Date/Time: 07/26/2018 19:52

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HABART,LAUREN,F  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: HABART,LAUREN  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 06/13/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19817080A3  
DRY GAS EXPIRY DATE: 09/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 20:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: SHAFFER

DATE/TIME OF TEST: 07/18/2018 21:37:47

---

Diagnostics Check	OK	21:29
Air Blank	0.000	21:29
Control Test	0.080	21:30
Air Blank	0.000	21:30
Subject Sample # 1	0.081	21:33
Breath Volume	1.500	21:33
Air Blank	0.000	21:34
Air Blank	0.000	21:35
Subject Sample # 2	0.086	21:36
Breath Volume	1.656	21:36
Air Blank	0.000	21:36
Control Test	0.077	21:37
Air Blank	0.000	21:37
Diagnostics Check	OK	21:37

---

Data Download Date/Time: 07/26/2018 19:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARRITY, THOMAS, A  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: GARRITY, THOMAS  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 06/13/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19817080A3  
DRY GAS EXPIRY DATE: 09/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: WEETHEE

DATE/TIME OF TEST: 07/20/2018 00:17:44

---

Diagnostics Check	OK	00:11
Air Blank	0.000	00:11
Control Test	0.080	00:12
Air Blank	0.000	00:12
Subject Sample # 1	0.112	00:13
Breath Volume	2.042	00:13
Air Blank	0.000	00:13
Air Blank	0.000	00:15
Subject Sample # 2	0.116	00:16
Breath Volume	2.945	00:16
Air Blank	0.000	00:16
Control Test	0.080	00:17
Air Blank	0.000	00:17
Diagnostics Check	OK	00:17

---

Data Download Date/Time: 07/26/2018 19:53



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ELBON,NICHOLAS,C  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: DRACH,STEFAN  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 06/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 04:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: CHAMBERLIN

DATE/TIME OF TEST: 07/01/2018 05:15:06

---

Diagnostics Check		OK	05:08
Air Blank		0.000	05:09
Control Test		0.082	05:09
Air Blank		0.000	05:10
Subject Sample	# 1	0.148	05:10
Breath Volume		2.609	05:10
Air Blank		0.000	05:11
Air Blank		0.000	05:13
Subject Sample	# 2	0.151	05:13
Breath Volume		2.421	05:13
Air Blank		0.000	05:14
Control Test		0.081	05:14
Air Blank		0.000	05:15
Diagnostics Check		OK	05:15

---

Data Download Date/Time: 07/05/2018 14:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: HARRIS,BRIAN  
ARRESTING OFFICER AGENCY: Ocoee PD  
LAST AGENCY INSPECTION: 06/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: DETRES MERCADO

DATE/TIME OF TEST: 07/04/2018 23:07:47

---

Diagnostics Check	OK	22:52	
Air Blank	0.000	22:52	
Control Test	0.082	22:52	
Air Blank	0.000	22:53	
Subject Sample	# 1 0.190	22:54	
Breath Volume	1.585	22:54	
Air Blank	0.000	22:55	
Air Blank	0.000	22:57	
Subject Sample	# 2 0.172	23:00	Volume Not Met
Breath Volume	1.023	23:00	
Air Blank	0.000	23:01	
Air Blank	0.000	23:02	
Subject Sample	# 3 0.138	23:06	Volume Not Met
Breath Volume	0.046	23:06	
Air Blank	0.000	23:06	
Control Test	0.081	23:07	
Air Blank	0.000	23:07	
Diagnostics Check	OK	23:07	

---

Data Download Date/Time: 07/05/2018 14:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ELBON,NICHOLAS,C  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: VINER,DENNIS  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 06/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: ALAMILLO ZOPIYACTLE

DATE/TIME OF TEST: 07/05/2018 00:14:54

---

Diagnostics Check	OK	00:08
Air Blank	0.000	00:09
Control Test	0.082	00:09
Air Blank	0.000	00:09
Subject Sample # 1	0.179	00:10
Breath Volume	2.484	00:10
Air Blank	0.000	00:11
Air Blank	0.000	00:12
Subject Sample # 2	0.182	00:13
Breath Volume	2.589	00:13
Air Blank	0.000	00:13
Control Test	0.081	00:14
Air Blank	0.000	00:14
Diagnostics Check	OK	00:14

---

Data Download Date/Time: 07/05/2018 14:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ELBON,NICHOLAS,C  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: AUSTIN,ADAM  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 06/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: JAMES

DATE/TIME OF TEST: 07/05/2018 03:03:52

---

Diagnostics Check	OK	02:57
Air Blank	0.000	02:58
Control Test	0.082	02:58
Air Blank	0.000	02:58
Subject Sample # 1	0.047	02:59
Breath Volume	3.472	02:59
Air Blank	0.000	03:00
Air Blank	0.000	03:01
Subject Sample # 2	0.042	03:02
Breath Volume	1.820	03:02
Air Blank	0.000	03:02
Control Test	0.081	03:03
Air Blank	0.000	03:03
Diagnostics Check	OK	03:03

---

Data Download Date/Time: 07/05/2018 14:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: HAWKINS,DEB  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 06/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: JAMES

DATE/TIME OF TEST: 07/05/2018 04:02:34

---

Diagnostics Check	OK	03:55
Air Blank	0.000	03:56
Control Test	0.082	03:56
Air Blank	0.000	03:57
Subject Sample # 1	0.134	03:57
Breath Volume	1.609	03:57
Air Blank	0.000	03:58
Air Blank	0.000	04:00
Subject Sample # 2	0.152	04:00
Breath Volume	3.250	04:00
Air Blank	0.000	04:01
Control Test	0.081	04:01
Air Blank	0.000	04:02
Diagnostics Check	OK	04:02

---

Data Download Date/Time: 07/05/2018 14:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ELBON,NICHOLAS,C  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: ALEXANDER,JARED  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 06/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 04:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:56

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 07/05/2018 05:41:47

---

Diagnostics Check		OK	05:25	
Air Blank		0.000	05:25	
Control Test		0.082	05:26	
Air Blank		0.000	05:26	
Subject Sample	# 1	0.153	05:29	Volume Not Met
Breath Volume		0.351	05:29	
Air Blank		0.000	05:30	
Air Blank		0.000	05:32	
Subject Sample	# 2	0.194	05:34	
Breath Volume		1.375	05:34	
Air Blank		0.000	05:35	
Air Blank		0.000	05:37	
Subject Sample	# 3	0.146	05:40	Volume Not Met
Breath Volume		0.167	05:40	
Air Blank		0.000	05:40	
Control Test		0.080	05:41	
Air Blank		0.000	05:41	
Diagnostics Check		OK	05:41	

---

Data Download Date/Time: 07/05/2018 14:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: DAVILA,SAMUEL  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 06/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: DESUE

DATE/TIME OF TEST: 07/08/2018 04:26:25

---

Diagnostics Check	OK	04:19
Air Blank	0.000	04:19
Control Test	0.083	04:20
Air Blank	0.000	04:20
Subject Sample # 1	0.142	04:21
Breath Volume	2.132	04:21
Air Blank	0.000	04:22
Air Blank	0.000	04:24
Subject Sample # 2	0.125	04:24
Breath Volume	1.113	04:24
Air Blank	0.000	04:25
Control Test	0.080	04:25
Air Blank	0.000	04:26
Diagnostics Check	OK	04:26

---

Data Download Date/Time: 07/11/2018 16:46

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ELBON,NICHOLAS,C  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: KNIGHT,RYAN  
ARRESTING OFFICER AGENCY: Maitland PD  
LAST AGENCY INSPECTION: 06/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: KLINKER

DATE/TIME OF TEST: 07/10/2018 04:20:46

---

Diagnostics Check	OK	04:14
Air Blank	0.000	04:15
Control Test	0.082	04:15
Air Blank	0.000	04:16
Subject Sample # 1	0.000	04:16
Breath Volume	1.113	04:16
Air Blank	0.000	04:17
Air Blank	0.000	04:19
Subject Sample # 2	0.000	04:19
Breath Volume	1.304	04:19
Air Blank	0.000	04:19
Control Test	0.083	04:20
Air Blank	0.000	04:20
Diagnostics Check	OK	04:20

---

Data Download Date/Time: 07/11/2018 16:46



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TIAMSON,DAWN,P  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: JOHNS,MARK  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 06/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: RAMOS RAMOS

DATE/TIME OF TEST: 07/13/2018 00:33:57

---

Diagnostics Check	OK	00:25
Air Blank	0.000	00:26
Control Test	0.083	00:26
Air Blank	0.000	00:27
Subject Sample # 1	0.135	00:28
Breath Volume	3.179	00:28
Air Blank	0.000	00:29
Air Blank	0.000	00:31
Subject Sample # 2	0.134	00:32
Breath Volume	2.031	00:32
Air Blank	0.000	00:33
Control Test	0.080	00:33
Air Blank	0.000	00:33
Diagnostics Check	OK	00:33

---

Data Download Date/Time: 07/19/2018 10:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: URIBE,SERGIO  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 06/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: VARGAS

DATE/TIME OF TEST: 07/14/2018 02:02:34

---

Diagnostics Check	OK	01:50	
Air Blank	0.000	01:51	
Control Test	0.082	01:51	
Air Blank	0.000	01:52	
Subject Sample	# 1 0.135	01:55	Volume Not Met
Breath Volume	0.234	01:55	
Air Blank	0.000	01:56	
Air Blank	0.000	01:57	
Subject Sample	# 2 0.125	02:00	Volume Not Met
Breath Volume	0.687	02:00	
Air Blank	0.000	02:01	
Control Test	0.080	02:01	
Air Blank	0.000	02:02	
Diagnostics Check	OK	02:02	

---

Data Download Date/Time: 07/19/2018 10:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: KREKEL,ALEC  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 06/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 07/14/2018 03:20:30

---

Diagnostics Check		OK	03:14
Air Blank		0.000	03:14
Control Test		0.081	03:14
Air Blank		0.000	03:15
Subject Sample	# 1	0.085	03:15
Breath Volume		2.242	03:15
Air Blank		0.000	03:16
Air Blank		0.000	03:18
Subject Sample	# 2	0.083	03:18
Breath Volume		2.132	03:18
Air Blank		0.000	03:19
Control Test		0.080	03:19
Air Blank		0.000	03:20
Diagnostics Check		OK	03:20

---

Data Download Date/Time: 07/19/2018 10:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: JORGE,JUNIOR  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 06/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 04:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: SAMAROO

DATE/TIME OF TEST: 07/14/2018 05:04:13

---

Diagnostics Check	OK	04:57
Air Blank	0.000	04:57
Control Test	0.082	04:57
Air Blank	0.000	04:58
Subject Sample # 1	0.140	04:59
Breath Volume	1.886	04:59
Air Blank	0.000	05:00
Air Blank	0.000	05:01
Subject Sample # 2	0.143	05:02
Breath Volume	4.285	05:02
Air Blank	0.000	05:03
Control Test	0.080	05:03
Air Blank	0.000	05:04
Diagnostics Check	OK	05:04

---

Data Download Date/Time: 07/19/2018 10:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARDEN,WILLIAM,L  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: SPENCER,JOHN  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 06/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 04:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:56

SUBJECT'S LAST NAME: NORRIS

DATE/TIME OF TEST: 07/14/2018 05:23:46

---

Diagnostics Check	OK	05:20	
Air Blank	0.000	05:20	
Control Test	0.080	05:21	
Air Blank	0.000	05:21	
Subject Sample	# 1	05:22	Subject Test Refused
Breath Volume	0.000	05:22	
Air Blank	0.000	05:22	
Control Test	0.081	05:23	
Air Blank	0.000	05:23	
Diagnostics Check	OK	05:23	

---

Data Download Date/Time: 07/19/2018 10:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: HAWKINS,DEB  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 06/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: PUGH

DATE/TIME OF TEST: 07/15/2018 04:17:11

---

Diagnostics Check	OK	04:10
Air Blank	0.000	04:11
Control Test	0.083	04:11
Air Blank	0.000	04:12
Subject Sample # 1	0.158	04:12
Breath Volume	1.957	04:12
Air Blank	0.000	04:13
Air Blank	0.000	04:15
Subject Sample # 2	0.163	04:15
Breath Volume	2.761	04:15
Air Blank	0.000	04:16
Control Test	0.081	04:16
Air Blank	0.000	04:17
Diagnostics Check	OK	04:17

---

Data Download Date/Time: 07/19/2018 10:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TIAMSON,DAWN,P  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: ANDREWS,GARY  
ARRESTING OFFICER AGENCY: Maitland PD  
LAST AGENCY INSPECTION: 07/19/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A1  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 20:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 07/20/2018 21:41:35

---

Diagnostics Check	OK	21:35
Air Blank	0.000	21:35
Control Test	0.082	21:36
Air Blank	0.000	21:36
Subject Sample # 1	0.122	21:37
Breath Volume	2.726	21:37
Air Blank	0.000	21:37
Air Blank	0.000	21:39
Subject Sample # 2	0.124	21:39
Breath Volume	2.367	21:39
Air Blank	0.000	21:40
Control Test	0.081	21:40
Air Blank	0.000	21:41
Diagnostics Check	OK	21:41

---

Data Download Date/Time: 07/25/2018 17:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: SIMPSON,JOSEPH  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 07/19/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A1  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: BAMBEI

DATE/TIME OF TEST: 07/21/2018 03:05:13

---

Diagnostics Check	OK	02:53
Air Blank	0.000	02:53
Control Test	0.082	02:53
Air Blank	0.000	02:54
Subject Sample # 1	0.138	02:57
Breath Volume	1.648	02:57
Air Blank	0.000	02:58
Air Blank	0.000	02:59
Subject Sample # 2	0.112	03:00
Breath Volume	1.949	03:00
Air Blank	0.000	03:01
Air Blank	0.000	03:03
Subject Sample # 3	0.118	03:03
Breath Volume	2.777	03:03
Air Blank	0.000	03:04
Control Test	0.078	03:04
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

---

Data Download Date/Time: 07/25/2018 17:36



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TIAMSON,DAWN,P  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: RAMIREZ,MATTHEW  
ARRESTING OFFICER AGENCY: Ocoee PD  
LAST AGENCY INSPECTION: 07/19/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A1  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: MONTES

DATE/TIME OF TEST: 07/21/2018 04:00:43

---

Diagnostics Check	OK	03:53
Air Blank	0.000	03:53
Control Test	0.082	03:53
Air Blank	0.000	03:54
Subject Sample # 1	0.191	03:55
Breath Volume	1.386	03:55
Air Blank	0.000	03:56
Air Blank	0.000	03:58
Subject Sample # 2	0.200	03:59
Breath Volume	2.277	03:59
Air Blank	0.000	03:59
Control Test	0.081	04:00
Air Blank	0.000	04:00
Diagnostics Check	OK	04:00

---

Data Download Date/Time: 07/25/2018 17:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMPSON,CALEB,N  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: NEWELL,JEFF  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 07/19/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A1  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 04:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: PALAEZOLIVEROS

DATE/TIME OF TEST: 07/21/2018 04:31:46

---

Diagnostics Check	OK	04:25
Air Blank	0.000	04:25
Control Test	0.081	04:26
Air Blank	0.000	04:26
Subject Sample # 1	0.161	04:27
Breath Volume	1.773	04:27
Air Blank	0.000	04:28
Air Blank	0.000	04:29
Subject Sample # 2	0.164	04:30
Breath Volume	2.660	04:30
Air Blank	0.000	04:30
Control Test	0.081	04:31
Air Blank	0.000	04:31
Diagnostics Check	OK	04:31

---

Data Download Date/Time: 07/25/2018 17:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLS,ARMAND,G  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: POWELL,KRISTI  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 07/19/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 15:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:23

SUBJECT'S LAST NAME: SOSA MORILLO

DATE/TIME OF TEST: 07/21/2018 17:04:23

---

Diagnostics Check	OK	16:58
Air Blank	0.000	16:58
Control Test	0.082	16:59
Air Blank	0.000	16:59
Subject Sample # 1	0.181	16:59
Breath Volume	1.265	16:59
Air Blank	0.000	17:00
Air Blank	0.000	17:02
Subject Sample # 2	0.171	17:02
Breath Volume	1.902	17:02
Air Blank	0.000	17:03
Control Test	0.082	17:03
Air Blank	0.000	17:04
Diagnostics Check	OK	17:04

---

Data Download Date/Time: 07/25/2018 17:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMPSON,CALEB,N  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: RAMIREZ,MATTHEW  
ARRESTING OFFICER AGENCY: Ocoee PD  
LAST AGENCY INSPECTION: 07/19/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A1  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 19:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: LOPEZGABRIEL

DATE/TIME OF TEST: 07/21/2018 20:52:52

---

Diagnostics Check	OK	20:45
Air Blank	0.000	20:45
Control Test	0.082	20:46
Air Blank	0.000	20:46
Subject Sample # 1	0.208	20:47
Breath Volume	1.332	20:47
Air Blank	0.000	20:48
Air Blank	0.000	20:49
Subject Sample # 2	0.222	20:51
Breath Volume	2.046	20:51
Air Blank	0.000	20:51
Control Test	0.081	20:52
Air Blank	0.000	20:52
Diagnostics Check	OK	20:52

---

Data Download Date/Time: 07/25/2018 17:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TIAMSON,DAWN,P  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: SPAIN,ALICIA  
ARRESTING OFFICER AGENCY: Winter Park PD  
LAST AGENCY INSPECTION: 07/19/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A1  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: RUIZPEREZ

DATE/TIME OF TEST: 07/21/2018 23:24:17

---

Diagnostics Check	OK	23:16
Air Blank	0.000	23:17
Control Test	0.082	23:17
Air Blank	0.000	23:18
Subject Sample # 1	0.208	23:18
Breath Volume	1.441	23:18
Air Blank	0.000	23:19
Air Blank	0.000	23:20
Subject Sample # 2	0.208	23:22
Breath Volume	1.625	23:22
Air Blank	0.000	23:23
Control Test	0.082	23:23
Air Blank	0.000	23:24
Diagnostics Check	OK	23:24

---

Data Download Date/Time: 07/25/2018 17:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: MARTINEZ,YAISEL  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 07/19/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A1  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: DIAZSANTOS

DATE/TIME OF TEST: 07/22/2018 02:37:10

---

Diagnostics Check		OK	02:30
Air Blank		0.000	02:30
Control Test		0.081	02:30
Air Blank		0.000	02:31
Subject Sample	# 1	0.137	02:32
Breath Volume		1.203	02:32
Air Blank		0.000	02:32
Air Blank		0.000	02:34
Subject Sample	# 2	0.140	02:35
Breath Volume		1.527	02:35
Air Blank		0.000	02:36
Control Test		0.080	02:36
Air Blank		0.000	02:37
Diagnostics Check		OK	02:37

---

Data Download Date/Time: 07/25/2018 17:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMPSON,CALEB,N  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: WILKES,DAVID  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 07/19/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A1  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:36  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: HOGARTH

DATE/TIME OF TEST: 07/22/2018 02:48:57

---

Diagnostics Check	OK	02:42
Air Blank	0.000	02:42
Control Test	0.081	02:43
Air Blank	0.000	02:43
Subject Sample # 1	0.194	02:44
Breath Volume	1.789	02:44
Air Blank	0.000	02:44
Air Blank	0.000	02:46
Subject Sample # 2	0.193	02:47
Breath Volume	1.335	02:47
Air Blank	0.000	02:48
Control Test	0.082	02:48
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

---

Data Download Date/Time: 07/25/2018 17:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TIAMSON,DAWN,P  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: HAWKINS,DEBORAH  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 07/19/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A1  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: FLORESBUENO

DATE/TIME OF TEST: 07/22/2018 04:45:33

---

Diagnostics Check	OK	04:39
Air Blank	0.000	04:39
Control Test	0.082	04:39
Air Blank	0.000	04:40
Subject Sample # 1	0.165	04:40
Breath Volume	3.875	04:40
Air Blank	0.000	04:41
Air Blank	0.000	04:43
Subject Sample # 2	0.169	04:43
Breath Volume	3.925	04:43
Air Blank	0.000	04:44
Control Test	0.081	04:44
Air Blank	0.000	04:45
Diagnostics Check	OK	04:45

---

Data Download Date/Time: 07/25/2018 17:36



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: SIMPSON,JOSEPH  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 07/19/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A1  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: DOWNS

DATE/TIME OF TEST: 07/23/2018 04:43:48

---

Diagnostics Check	OK	04:36
Air Blank	0.000	04:36
Control Test	0.082	04:36
Air Blank	0.000	04:37
Subject Sample # 1	0.190	04:38
Breath Volume	1.863	04:38
Air Blank	0.000	04:39
Air Blank	0.000	04:41
Subject Sample # 2	0.206	04:42
Breath Volume	1.316	04:42
Air Blank	0.000	04:42
Control Test	0.080	04:43
Air Blank	0.000	04:43
Diagnostics Check	OK	04:43

---

Data Download Date/Time: 07/25/2018 17:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: KREKEL,ALEC  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 07/19/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A1  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: ACOSTA

DATE/TIME OF TEST: 07/24/2018 01:50:14

---

Diagnostics Check	OK	01:44
Air Blank	0.000	01:44
Control Test	0.083	01:44
Air Blank	0.000	01:45
Subject Sample # 1	0.103	01:45
Breath Volume	2.164	01:45
Air Blank	0.000	01:46
Air Blank	0.000	01:48
Subject Sample # 2	0.103	01:48
Breath Volume	2.464	01:48
Air Blank	0.000	01:49
Control Test	0.081	01:49
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

---

Data Download Date/Time: 07/25/2018 17:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: LOWE,KEVIN  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 07/19/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:01  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 07/26/2018 01:13:52

---

Diagnostics Check	OK	01:07
Air Blank	0.000	01:07
Control Test	0.082	01:08
Air Blank	0.000	01:08
Subject Sample # 1	0.226	01:08
Breath Volume	1.144	01:08
Air Blank	0.000	01:09
Air Blank	0.000	01:11
Subject Sample # 2	0.230	01:12
Breath Volume	1.554	01:12
Air Blank	0.000	01:12
Control Test	0.081	01:13
Air Blank	0.000	01:13
Diagnostics Check	OK	01:13

---

Data Download Date/Time: 08/14/2018 11:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMPSON,CALEB,N  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: KEANE,PETER  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 07/19/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:51  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: FLORES

DATE/TIME OF TEST: 07/26/2018 22:57:50

---

Diagnostics Check	OK	22:51
Air Blank	0.000	22:51
Control Test	0.083	22:51
Air Blank	0.000	22:52
Subject Sample # 1	0.182	22:52
Breath Volume	2.683	22:52
Air Blank	0.000	22:53
Air Blank	0.000	22:55
Subject Sample # 2	0.176	22:56
Breath Volume	3.203	22:56
Air Blank	0.000	22:56
Control Test	0.081	22:57
Air Blank	0.000	22:57
Diagnostics Check	OK	22:57

---

Data Download Date/Time: 08/14/2018 11:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: VENTURA,CHRISTOPHER  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 07/19/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 07/27/2018 01:12:56

---

Diagnostics Check	OK	01:05
Air Blank	0.000	01:05
Control Test	0.082	01:06
Air Blank	0.000	01:06
Subject Sample # 1	0.158	01:08
Breath Volume	1.867	01:08
Air Blank	0.000	01:09
Air Blank	0.000	01:10
Subject Sample # 2	0.159	01:11
Breath Volume	2.031	01:11
Air Blank	0.000	01:12
Control Test	0.079	01:12
Air Blank	0.000	01:12
Diagnostics Check	OK	01:12

---

Data Download Date/Time: 08/14/2018 11:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMPSON,CALEB,N  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: MARTINEZ,ERNESTO  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 07/19/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: CHADA

DATE/TIME OF TEST: 07/27/2018 02:54:44

---

Diagnostics Check		OK	02:47
Air Blank		0.000	02:48
Control Test		0.083	02:48
Air Blank		0.000	02:49
Subject Sample	# 1	0.214	02:50
Breath Volume		4.382	02:50
Air Blank		0.000	02:50
Air Blank		0.000	02:52
Subject Sample	# 2	0.200	02:53
Breath Volume		2.496	02:53
Air Blank		0.000	02:53
Control Test		0.080	02:54
Air Blank		0.000	02:54
Diagnostics Check		OK	02:54

---

Data Download Date/Time: 08/14/2018 11:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TIAMSON,DAWN,P  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: KREKEL,ALEC  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 07/19/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: MORTON

DATE/TIME OF TEST: 07/27/2018 04:10:41

---

Diagnostics Check	OK	03:59	
Air Blank	0.000	03:59	
Control Test	0.082	04:00	
Air Blank	0.000	04:00	
Subject Sample	# 1	04:03	No Sample Provided
Breath Volume	0.000	04:03	
Air Blank	0.000	04:04	
Air Blank	0.000	04:06	
Subject Sample	# 2	04:09	No Sample Provided
Breath Volume	0.000	04:09	
Air Blank	0.000	04:09	
Control Test	0.082	04:10	
Air Blank	0.000	04:10	
Diagnostics Check	OK	04:10	

---

Data Download Date/Time: 08/14/2018 11:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DENSON,TRAVIS,W  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: SPENCER,JOHN  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 07/19/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: AGOSTO ROSADO

DATE/TIME OF TEST: 07/28/2018 01:31:25

---

Diagnostics Check		OK	01:24
Air Blank		0.000	01:25
Control Test		0.082	01:25
Air Blank		0.000	01:25
Subject Sample	# 1	0.210	01:26
Breath Volume		4.378	01:26
Air Blank		0.000	01:27
Air Blank		0.000	01:29
Subject Sample	# 2	0.198	01:29
Breath Volume		3.601	01:29
Air Blank		0.000	01:30
Control Test		0.080	01:30
Air Blank		0.000	01:31
Diagnostics Check		OK	01:31

---

Data Download Date/Time: 08/14/2018 11:53



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TIAMSON,DAWN,P  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: BORRAS,TAMARA  
ARRESTING OFFICER AGENCY: Winter Garden PD  
LAST AGENCY INSPECTION: 07/19/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: ROLONGONZALEZ

DATE/TIME OF TEST: 07/28/2018 01:51:22

---

Diagnostics Check	OK	01:42
Air Blank	0.000	01:42
Control Test	0.080	01:42
Air Blank	0.000	01:43
Subject Sample # 1	0.162	01:46
Breath Volume	1.406	01:46
Air Blank	0.000	01:47
Air Blank	0.000	01:48
Subject Sample # 2	0.175	01:49
Breath Volume	1.429	01:49
Air Blank	0.000	01:50
Control Test	0.079	01:50
Air Blank	0.000	01:51
Diagnostics Check	OK	01:51

---

Data Download Date/Time: 08/14/2018 11:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: GONZALEZ,ELVIN  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 07/19/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: GOMEZCASILLAS

DATE/TIME OF TEST: 07/28/2018 03:15:06

---

Diagnostics Check	OK	03:08
Air Blank	0.000	03:09
Control Test	0.083	03:09
Air Blank	0.000	03:10
Subject Sample # 1	0.111	03:10
Breath Volume	3.773	03:10
Air Blank	0.000	03:11
Air Blank	0.000	03:13
Subject Sample # 2	0.108	03:13
Breath Volume	3.582	03:13
Air Blank	0.000	03:14
Control Test	0.081	03:14
Air Blank	0.000	03:15
Diagnostics Check	OK	03:15

---

Data Download Date/Time: 08/14/2018 11:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMPSON,CALEB,N  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: CORREA,ASHLEY  
ARRESTING OFFICER AGENCY: Ocoee PD  
LAST AGENCY INSPECTION: 07/19/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:06

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 07/28/2018 22:38:31

---

Diagnostics Check	OK	22:29
Air Blank	0.000	22:30
Control Test	0.082	22:30
Air Blank	0.000	22:31
Subject Sample # 1	0.199	22:33
Breath Volume	3.386	22:33
Air Blank	0.000	22:34
Air Blank	0.000	22:36
Subject Sample # 2	0.186	22:36
Breath Volume	2.121	22:36
Air Blank	0.000	22:37
Control Test	0.079	22:37
Air Blank	0.000	22:38
Diagnostics Check	OK	22:38

---

Data Download Date/Time: 08/14/2018 11:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TIAMSON,DAWN,P  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: HAM,ALEXANDER  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 07/19/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: PRADO

DATE/TIME OF TEST: 07/28/2018 23:19:28

---

Diagnostics Check	OK	23:11
Air Blank	0.000	23:12
Control Test	0.082	23:12
Air Blank	0.000	23:13
Subject Sample # 1	0.171	23:14
Breath Volume	1.335	23:14
Air Blank	0.000	23:14
Air Blank	0.000	23:16
Subject Sample # 2	0.178	23:17
Breath Volume	3.593	23:17
Air Blank	0.000	23:18
Control Test	0.080	23:18
Air Blank	0.000	23:19
Diagnostics Check	OK	23:19

---

Data Download Date/Time: 08/14/2018 11:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: BUSCHOR,JOSEPH  
ARRESTING OFFICER AGENCY: FHP Troop K  
LAST AGENCY INSPECTION: 07/19/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: DAILEY

DATE/TIME OF TEST: 07/29/2018 03:43:49

---

Diagnostics Check	OK	03:32	
Air Blank	0.000	03:32	
Control Test	0.081	03:33	
Air Blank	0.000	03:33	
Subject Sample	# 1	03:36	No Sample Provided
Breath Volume	0.000	03:36	
Air Blank	0.000	03:37	
Air Blank	0.000	03:39	
Subject Sample	# 2	03:42	No Sample Provided
Breath Volume	0.000	03:42	
Air Blank	0.000	03:42	
Control Test	0.082	03:43	
Air Blank	0.000	03:43	
Diagnostics Check	OK	03:43	

---

Data Download Date/Time: 08/14/2018 11:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: SPENCER,JOHN  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 07/19/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:46  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: CRAWFORD

DATE/TIME OF TEST: 07/29/2018 04:37:47

---

Diagnostics Check	OK	04:31
Air Blank	0.000	04:31
Control Test	0.081	04:31
Air Blank	0.000	04:32
Subject Sample # 1	0.141	04:33
Breath Volume	3.515	04:33
Air Blank	0.000	04:33
Air Blank	0.000	04:35
Subject Sample # 2	0.138	04:36
Breath Volume	3.312	04:36
Air Blank	0.000	04:36
Control Test	0.081	04:37
Air Blank	0.000	04:37
Diagnostics Check	OK	04:37

---

Data Download Date/Time: 08/14/2018 11:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: PEREZ,JUAN  
ARRESTING OFFICER AGENCY: FHP Troop K  
LAST AGENCY INSPECTION: 07/19/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 04:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:07

SUBJECT'S LAST NAME: BRISSEAU

DATE/TIME OF TEST: 07/30/2018 05:40:20

---

Diagnostics Check		OK	05:30
Air Blank		0.000	05:30
Control Test		0.083	05:31
Air Blank		0.000	05:31
Subject Sample	# 1	0.090	05:34
Breath Volume		1.226	05:34
Air Blank		0.000	05:35
Air Blank		0.000	05:36
Subject Sample	# 2	0.077	05:38
Breath Volume		1.179	05:38
Air Blank		0.000	05:39
Control Test		0.080	05:39
Air Blank		0.000	05:40
Diagnostics Check		OK	05:40

---

Data Download Date/Time: 08/14/2018 11:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: WING,JENNIFER  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 07/19/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 15:04  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 15:36

SUBJECT'S LAST NAME: HOLTZ

DATE/TIME OF TEST: 07/30/2018 16:06:23

---

Diagnostics Check	OK	15:59
Air Blank	0.000	16:00
Control Test	0.082	16:00
Air Blank	0.000	16:01
Subject Sample # 1	0.271	16:01
Breath Volume	2.035	16:01
Air Blank	0.000	16:02
Air Blank	0.000	16:04
Subject Sample # 2	0.283	16:04
Breath Volume	3.152	16:04
Air Blank	0.000	16:05
Control Test	0.081	16:05
Air Blank	0.000	16:06
Diagnostics Check	OK	16:06

---

Data Download Date/Time: 08/14/2018 11:53



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: CANADA,JONATHAN  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 07/19/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 06:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:41

SUBJECT'S LAST NAME: GRIEK ROBERTSON

DATE/TIME OF TEST: 07/31/2018 07:13:32

---

Diagnostics Check	OK	07:09	
Air Blank	0.000	07:09	
Control Test	0.082	07:10	
Air Blank	0.000	07:10	
Subject Sample	# 1	07:12	Subject Test Refused
Breath Volume	0.000	07:12	
Air Blank	0.000	07:12	
Control Test	0.082	07:12	
Air Blank	0.000	07:13	
Diagnostics Check	OK	07:13	

---

Data Download Date/Time: 08/14/2018 11:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: DRACH, STEFAN  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 06/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: ALMENGOL

DATE/TIME OF TEST: 07/05/2018 01:23:13

---

Diagnostics Check		OK	01:17
Air Blank		0.000	01:17
Control Test		0.081	01:17
Air Blank		0.000	01:18
Subject Sample	# 1	0.193	01:18
Breath Volume		1.761	01:18
Air Blank		0.000	01:19
Air Blank		0.000	01:21
Subject Sample	# 2	0.187	01:21
Breath Volume		1.957	01:21
Air Blank		0.000	01:22
Control Test		0.080	01:22
Air Blank		0.000	01:23
Diagnostics Check		OK	01:23

---

Data Download Date/Time: 07/05/2018 14:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: KERN, ETHAN  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 06/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:11  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: BRYCE

DATE/TIME OF TEST: 07/05/2018 03:10:32

---

Diagnostics Check		OK	02:59	
Air Blank		0.000	02:59	
Control Test		0.081	02:59	
Air Blank		0.000	03:00	
Subject Sample	# 1	0.208	03:03	Volume Not Met
Breath Volume		0.261	03:03	
Air Blank		0.000	03:04	
Air Blank		0.000	03:06	
Subject Sample	# 2		03:09	No Sample Provided
Breath Volume		0.000	03:09	
Air Blank		0.000	03:09	
Control Test		0.080	03:09	
Air Blank		0.000	03:10	
Diagnostics Check		OK	03:10	

---

Data Download Date/Time: 07/05/2018 14:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: ELLER, ASHLEY  
ARRESTING OFFICER AGENCY: Apopka PD  
LAST AGENCY INSPECTION: 06/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: ZEALOR

DATE/TIME OF TEST: 07/07/2018 02:31:49

---

Diagnostics Check	OK	02:25
Air Blank	0.000	02:26
Control Test	0.080	02:26
Air Blank	0.000	02:26
Subject Sample # 1	0.190	02:27
Breath Volume	2.460	02:27
Air Blank	0.000	02:27
Air Blank	0.000	02:29
Subject Sample # 2	0.187	02:30
Breath Volume	1.171	02:30
Air Blank	0.000	02:30
Control Test	0.080	02:31
Air Blank	0.000	02:31
Diagnostics Check	OK	02:31

---

Data Download Date/Time: 07/11/2018 16:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: PEERS,MICHAEL  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 06/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 04:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: NEGRON SANTOS

DATE/TIME OF TEST: 07/07/2018 05:16:01

---

Diagnostics Check		OK	05:09
Air Blank		0.000	05:10
Control Test		0.080	05:10
Air Blank		0.000	05:11
Subject Sample	# 1	0.231	05:11
Breath Volume		1.875	05:11
Air Blank		0.000	05:12
Air Blank		0.000	05:14
Subject Sample	# 2	0.225	05:14
Breath Volume		2.003	05:14
Air Blank		0.000	05:15
Control Test		0.081	05:15
Air Blank		0.000	05:15
Diagnostics Check		OK	05:16

---

Data Download Date/Time: 07/11/2018 16:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: CROCK, NICOLLE  
ARRESTING OFFICER AGENCY: Edgewood PD  
LAST AGENCY INSPECTION: 06/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: FIGUEROA RODRIGUEZ

DATE/TIME OF TEST: 07/18/2018 04:02:43

---

Diagnostics Check	OK	03:56
Air Blank	0.000	03:56
Control Test	0.081	03:57
Air Blank	0.000	03:57
Subject Sample # 1	0.111	03:58
Breath Volume	2.593	03:58
Air Blank	0.000	03:58
Air Blank	0.000	04:00
Subject Sample # 2	0.105	04:01
Breath Volume	2.953	04:01
Air Blank	0.000	04:01
Control Test	0.080	04:02
Air Blank	0.000	04:02
Diagnostics Check	OK	04:02

---

Data Download Date/Time: 07/19/2018 10:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: HANSTEIN, AUSTIN  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 06/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: RIOSCASTRO

DATE/TIME OF TEST: 07/19/2018 01:24:59

---

Diagnostics Check		OK	01:13	
Air Blank		0.000	01:13	
Control Test		0.081	01:14	
Air Blank		0.000	01:14	
Subject Sample	# 1	0.218	01:17	Volume Not Met
Breath Volume		0.230	01:17	
Air Blank		0.000	01:18	
Air Blank		0.000	01:20	
Subject Sample	# 2	0.199	01:23	Volume Not Met
Breath Volume		0.332	01:23	
Air Blank		0.000	01:24	
Control Test		0.081	01:24	
Air Blank		0.000	01:24	
Diagnostics Check		OK	01:24	

---

Data Download Date/Time: 07/19/2018 10:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARDEN,WILLIAM,L  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: PEERS,MICHAEL  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 07/19/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: CANTRELL

DATE/TIME OF TEST: 07/23/2018 01:11:34

---

Diagnostics Check		OK	01:05
Air Blank		0.000	01:05
Control Test		0.081	01:06
Air Blank		0.000	01:06
Subject Sample	# 1	0.145	01:07
Breath Volume		2.289	01:07
Air Blank		0.000	01:07
Air Blank		0.000	01:09
Subject Sample	# 2	0.132	01:10
Breath Volume		1.710	01:10
Air Blank		0.000	01:10
Control Test		0.080	01:10
Air Blank		0.000	01:11
Diagnostics Check		OK	01:11

---

Data Download Date/Time: 07/25/2018 17:16



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARDEN, WILLIAM, L  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: GERARDI, GREGORY  
ARRESTING OFFICER AGENCY: Winter Park PD  
LAST AGENCY INSPECTION: 07/19/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: GAHAGAN

DATE/TIME OF TEST: 07/24/2018 01:50:18

---

Diagnostics Check	OK	01:41	
Air Blank	0.000	01:42	
Control Test	0.081	01:42	
Air Blank	0.000	01:42	
Subject Sample	# 1	01:45	No Sample Provided
Breath Volume	0.000	01:45	
Air Blank	0.000	01:46	
Air Blank	0.000	01:48	
Subject Sample	# 2	01:48	Subject Test Refused
Breath Volume	0.000	01:48	
Air Blank	0.000	01:49	
Control Test	0.080	01:49	
Air Blank	0.000	01:50	
Diagnostics Check	OK	01:50	

---

Data Download Date/Time: 07/25/2018 17:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARDEN,WILLIAM,L  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: MORALES,JOSE  
ARRESTING OFFICER AGENCY: Central FL Univ PD  
LAST AGENCY INSPECTION: 07/19/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 07/24/2018 02:56:03

---

Diagnostics Check	OK	02:49
Air Blank	0.000	02:49
Control Test	0.081	02:50
Air Blank	0.000	02:50
Subject Sample # 1	0.166	02:51
Breath Volume	1.695	02:51
Air Blank	0.000	02:52
Air Blank	0.000	02:53
Subject Sample # 2	0.165	02:54
Breath Volume	1.945	02:54
Air Blank	0.000	02:55
Control Test	0.081	02:55
Air Blank	0.000	02:55
Diagnostics Check	OK	02:56

---

Data Download Date/Time: 07/25/2018 17:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARDEN,WILLIAM,L  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: HERNANDEZ,YASSELL  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 07/19/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 20:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: DAVILA

DATE/TIME OF TEST: 07/25/2018 21:08:16

---

Diagnostics Check	OK	20:59	
Air Blank	0.000	21:00	
Control Test	0.081	21:00	
Air Blank	0.000	21:01	
Subject Sample	# 1	21:04	No Sample Provided
Breath Volume	0.000	21:04	
Air Blank	0.000	21:04	
Air Blank	0.000	21:06	
Subject Sample	# 2	21:06	Subject Test Refused
Breath Volume	0.000	21:06	
Air Blank	0.000	21:07	
Control Test	0.082	21:07	
Air Blank	0.000	21:08	
Diagnostics Check	OK	21:08	

---

Data Download Date/Time: 08/14/2018 07:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: MORALES,JOSE  
ARRESTING OFFICER AGENCY: Central FL Univ PD  
LAST AGENCY INSPECTION: 06/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A1  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: SORIAL

DATE/TIME OF TEST: 07/01/2018 03:13:59

---

Diagnostics Check	OK	03:06
Air Blank	0.000	03:07
Control Test	0.079	03:07
Air Blank	0.000	03:08
Subject Sample # 1	0.103	03:09
Breath Volume	1.964	03:09
Air Blank	0.000	03:09
Air Blank	0.000	03:11
Subject Sample # 2	0.099	03:12
Breath Volume	3.421	03:12
Air Blank	0.000	03:13
Control Test	0.080	03:13
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

---

Data Download Date/Time: 07/05/2018 14:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ, MICHAEL, J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: NUNEZ, ANDREW  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 06/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A1  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 08:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:49

SUBJECT'S LAST NAME: PAREDESRODRIGUEZ

DATE/TIME OF TEST: 07/02/2018 09:20:54

---

Diagnostics Check	OK	09:14
Air Blank	0.000	09:15
Control Test	0.080	09:15
Air Blank	0.000	09:15
Subject Sample # 1	0.043	09:16
Breath Volume	1.820	09:16
Air Blank	0.000	09:17
Air Blank	0.000	09:18
Subject Sample # 2	0.042	09:19
Breath Volume	1.796	09:19
Air Blank	0.000	09:19
Control Test	0.080	09:20
Air Blank	0.000	09:20
Diagnostics Check	OK	09:20

---

Data Download Date/Time: 07/05/2018 14:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TIAMSON,DAWN,P  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: ROBBINS,NATHAN  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 06/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A1  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 03:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: SAMMY

DATE/TIME OF TEST: 07/04/2018 04:12:39

---

Diagnostics Check	OK	04:01	
Air Blank	0.000	04:01	
Control Test	0.079	04:02	
Air Blank	0.000	04:02	
Subject Sample	# 1	04:05	No Sample Provided
Breath Volume	0.000	04:05	
Air Blank	0.000	04:06	
Air Blank	0.000	04:08	
Subject Sample	# 2	04:11	No Sample Provided
Breath Volume	0.000	04:11	
Air Blank	0.000	04:11	
Control Test	0.080	04:12	
Air Blank	0.000	04:12	
Diagnostics Check	OK	04:12	

---

Data Download Date/Time: 07/05/2018 14:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMPSON,CALEB,N  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: POPP,ADAM  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 06/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A1  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: BUI  
ARREST TIME: 20:46  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:19

SUBJECT'S LAST NAME: MORENO-RIVERA

DATE/TIME OF TEST: 07/04/2018 21:55:54

---

Diagnostics Check	OK	21:44	
Air Blank	0.000	21:44	
Control Test	0.081	21:45	
Air Blank	0.000	21:45	
Subject Sample	# 1	21:48	No Sample Provided
Breath Volume	0.000	21:48	
Air Blank	0.000	21:49	
Air Blank	0.000	21:51	
Subject Sample	# 2	21:54	No Sample Provided
Breath Volume	0.000	21:54	
Air Blank	0.000	21:54	
Control Test	0.080	21:55	
Air Blank	0.000	21:55	
Diagnostics Check	OK	21:55	

---

Data Download Date/Time: 07/05/2018 14:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: MORALES,JOSE  
ARRESTING OFFICER AGENCY: Central FL Univ PD  
LAST AGENCY INSPECTION: 06/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A1  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 23:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: VALENCIA

DATE/TIME OF TEST: 07/05/2018 00:20:01

---

Diagnostics Check	OK	00:12
Air Blank	0.000	00:13
Control Test	0.080	00:13
Air Blank	0.000	00:13
Subject Sample # 1	0.105	00:14
Breath Volume	1.347	00:14
Air Blank	0.000	00:15
Air Blank	0.000	00:17
Subject Sample # 2	0.113	00:18
Breath Volume	2.683	00:18
Air Blank	0.000	00:19
Control Test	0.080	00:19
Air Blank	0.000	00:19
Diagnostics Check	OK	00:20

---

Data Download Date/Time: 07/05/2018 14:39



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMPSON,CALEB,N  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: HARRIS,BRIAN  
ARRESTING OFFICER AGENCY: Ocoee PD  
LAST AGENCY INSPECTION: 06/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A1  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 00:47  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 07/05/2018 01:38:51

---

Diagnostics Check	OK	01:32
Air Blank	0.000	01:33
Control Test	0.080	01:33
Air Blank	0.000	01:33
Subject Sample # 1	0.183	01:34
Breath Volume	1.582	01:34
Air Blank	0.000	01:34
Air Blank	0.000	01:36
Subject Sample # 2	0.182	01:37
Breath Volume	1.648	01:37
Air Blank	0.000	01:37
Control Test	0.079	01:38
Air Blank	0.000	01:38
Diagnostics Check	OK	01:38

---

Data Download Date/Time: 07/05/2018 14:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: DONAKOWSKI,JUSTIN  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 06/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A1  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: JIMENEZ GOMEZ

DATE/TIME OF TEST: 07/05/2018 02:58:37

---

Diagnostics Check	OK	02:52
Air Blank	0.000	02:52
Control Test	0.080	02:53
Air Blank	0.000	02:53
Subject Sample # 1	0.211	02:53
Breath Volume	3.144	02:53
Air Blank	0.000	02:54
Air Blank	0.000	02:56
Subject Sample # 2	0.219	02:57
Breath Volume	3.605	02:57
Air Blank	0.000	02:57
Control Test	0.078	02:58
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

---

Data Download Date/Time: 07/05/2018 14:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: IRELAND,DAVID  
ARRESTING OFFICER AGENCY: Edgewood PD  
LAST AGENCY INSPECTION: 06/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A1  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: BONAMY

DATE/TIME OF TEST: 07/05/2018 03:39:52

---

Diagnostics Check		OK	03:30
Air Blank		0.000	03:30
Control Test		0.079	03:30
Air Blank		0.000	03:31
Subject Sample	# 1	0.105	03:34
Breath Volume		2.640	03:34
Air Blank		0.000	03:34
Air Blank		0.000	03:36
Subject Sample	# 2	0.106	03:38
Breath Volume		2.375	03:38
Air Blank		0.000	03:38
Control Test		0.080	03:39
Air Blank		0.000	03:39
Diagnostics Check		OK	03:39

---

Data Download Date/Time: 07/05/2018 14:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMPSON,CALEB,N  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: MCWHERTER,PETER  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 06/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A1  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 17:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:23

SUBJECT'S LAST NAME: CROSBY

DATE/TIME OF TEST: 07/05/2018 18:52:16

---

Diagnostics Check	OK	18:45
Air Blank	0.000	18:46
Control Test	0.079	18:46
Air Blank	0.000	18:47
Subject Sample # 1	0.000	18:47
Breath Volume	1.937	18:47
Air Blank	0.000	18:48
Air Blank	0.000	18:50
Subject Sample # 2	0.000	18:50
Breath Volume	2.054	18:50
Air Blank	0.000	18:51
Control Test	0.080	18:51
Air Blank	0.000	18:52
Diagnostics Check	OK	18:52

---

Data Download Date/Time: 07/11/2018 16:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: SOUD,AHMAD  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 06/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A1  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 00:54  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: VARGAS

DATE/TIME OF TEST: 07/06/2018 01:42:26

---

Diagnostics Check	OK	01:36
Air Blank	0.000	01:36
Control Test	0.080	01:37
Air Blank	0.000	01:37
Subject Sample # 1	0.177	01:38
Breath Volume	2.105	01:38
Air Blank	0.000	01:38
Air Blank	0.000	01:40
Subject Sample # 2	0.173	01:40
Breath Volume	2.437	01:40
Air Blank	0.000	01:41
Control Test	0.080	01:41
Air Blank	0.000	01:42
Diagnostics Check	OK	01:42

---

Data Download Date/Time: 07/11/2018 16:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: AUSTIN,ADAM  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 06/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A1  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 04:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: RUEDA

DATE/TIME OF TEST: 07/06/2018 05:21:28

---

Diagnostics Check		OK	05:15
Air Blank		0.000	05:15
Control Test		0.081	05:15
Air Blank		0.000	05:16
Subject Sample	# 1	0.068	05:17
Breath Volume		2.062	05:17
Air Blank		0.000	05:17
Air Blank		0.000	05:19
Subject Sample	# 2	0.068	05:19
Breath Volume		2.085	05:19
Air Blank		0.000	05:20
Control Test		0.080	05:20
Air Blank		0.000	05:21
Diagnostics Check		OK	05:21

---

Data Download Date/Time: 07/11/2018 16:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TIAMSON,DAWN,P  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: WEAVER,WILLIAM  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 06/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A1  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 23:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 07/07/2018 00:58:49

---

Diagnostics Check	OK	00:51
Air Blank	0.000	00:52
Control Test	0.079	00:52
Air Blank	0.000	00:53
Subject Sample # 1	0.182	00:54
Breath Volume	2.000	00:54
Air Blank	0.000	00:54
Air Blank	0.000	00:56
Subject Sample # 2	0.183	00:57
Breath Volume	1.550	00:57
Air Blank	0.000	00:57
Control Test	0.079	00:58
Air Blank	0.000	00:58
Diagnostics Check	OK	00:58

---

Data Download Date/Time: 07/11/2018 16:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: OCAMPO,OSCAR  
ARRESTING OFFICER AGENCY: Maitland PD  
LAST AGENCY INSPECTION: 06/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A1  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 00:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: SALVANT

DATE/TIME OF TEST: 07/07/2018 01:54:42

---

Diagnostics Check	OK	01:48
Air Blank	0.000	01:48
Control Test	0.079	01:49
Air Blank	0.000	01:49
Subject Sample # 1	0.147	01:50
Breath Volume	2.859	01:50
Air Blank	0.000	01:50
Air Blank	0.000	01:52
Subject Sample # 2	0.146	01:53
Breath Volume	1.242	01:53
Air Blank	0.000	01:53
Control Test	0.079	01:54
Air Blank	0.000	01:54
Diagnostics Check	OK	01:54

---

Data Download Date/Time: 07/11/2018 16:30



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: COOP,JOHN  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 06/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A1  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: COLSON

DATE/TIME OF TEST: 07/07/2018 02:44:59

---

Diagnostics Check	OK	02:38
Air Blank	0.000	02:39
Control Test	0.080	02:39
Air Blank	0.000	02:40
Subject Sample # 1	0.199	02:40
Breath Volume	3.406	02:40
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.193	02:43
Breath Volume	2.039	02:43
Air Blank	0.000	02:44
Control Test	0.079	02:44
Air Blank	0.000	02:44
Diagnostics Check	OK	02:44

---

Data Download Date/Time: 07/11/2018 16:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TIAMSON,DAWN,P  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: HARRIS,BRIAN  
ARRESTING OFFICER AGENCY: Ocoee PD  
LAST AGENCY INSPECTION: 06/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A1  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: RICH

DATE/TIME OF TEST: 07/07/2018 03:25:02

---

Diagnostics Check	OK	03:18
Air Blank	0.000	03:18
Control Test	0.080	03:19
Air Blank	0.000	03:19
Subject Sample # 1	0.122	03:20
Breath Volume	1.671	03:20
Air Blank	0.000	03:20
Air Blank	0.000	03:22
Subject Sample # 2	0.116	03:23
Breath Volume	2.160	03:23
Air Blank	0.000	03:24
Control Test	0.079	03:24
Air Blank	0.000	03:24
Diagnostics Check	OK	03:25

---

Data Download Date/Time: 07/11/2018 16:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: HANSTINE,AUSTIN  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 06/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A1  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: BUSCH

DATE/TIME OF TEST: 07/07/2018 04:14:47

---

Diagnostics Check	OK	04:08
Air Blank	0.000	04:08
Control Test	0.079	04:09
Air Blank	0.000	04:09
Subject Sample # 1	0.228	04:10
Breath Volume	3.218	04:10
Air Blank	0.000	04:10
Air Blank	0.000	04:12
Subject Sample # 2	0.224	04:13
Breath Volume	3.871	04:13
Air Blank	0.000	04:13
Control Test	0.080	04:14
Air Blank	0.000	04:14
Diagnostics Check	OK	04:14

---

Data Download Date/Time: 07/11/2018 16:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ, MICHAEL, J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: BELARDES, DAVID  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 06/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A1  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 07:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:05

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 07/07/2018 08:39:34

---

Diagnostics Check	OK	08:33
Air Blank	0.000	08:33
Control Test	0.080	08:34
Air Blank	0.000	08:34
Subject Sample # 1	0.175	08:35
Breath Volume	1.433	08:35
Air Blank	0.000	08:35
Air Blank	0.000	08:37
Subject Sample # 2	0.179	08:38
Breath Volume	1.523	08:38
Air Blank	0.000	08:38
Control Test	0.080	08:38
Air Blank	0.000	08:39
Diagnostics Check	OK	08:39

---

Data Download Date/Time: 07/11/2018 16:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TIAMSON,DAWN,P  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: WEAVER,WILLIAM  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 06/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A1  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 23:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: JULIUS

DATE/TIME OF TEST: 07/08/2018 00:11:53

---

Diagnostics Check	OK	00:05
Air Blank	0.000	00:06
Control Test	0.080	00:06
Air Blank	0.000	00:07
Subject Sample # 1	0.118	00:07
Breath Volume	2.082	00:07
Air Blank	0.000	00:08
Air Blank	0.000	00:09
Subject Sample # 2	0.113	00:10
Breath Volume	1.710	00:10
Air Blank	0.000	00:10
Control Test	0.080	00:11
Air Blank	0.000	00:11
Diagnostics Check	OK	00:11

---

Data Download Date/Time: 07/11/2018 16:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: VAN SHOONHOVEN,JUSTIN  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 06/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A1  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 03:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: BARRETO RIVERA

DATE/TIME OF TEST: 07/08/2018 03:56:46

---

Diagnostics Check	OK	03:50
Air Blank	0.000	03:50
Control Test	0.080	03:50
Air Blank	0.000	03:51
Subject Sample # 1	0.151	03:52
Breath Volume	2.230	03:52
Air Blank	0.000	03:52
Air Blank	0.000	03:54
Subject Sample # 2	0.161	03:55
Breath Volume	3.558	03:55
Air Blank	0.000	03:55
Control Test	0.080	03:56
Air Blank	0.000	03:56
Diagnostics Check	OK	03:56

---

Data Download Date/Time: 07/11/2018 16:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DENSON, TRAVIS, W  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: QUARTERMAN, TAHVANI  
ARRESTING OFFICER AGENCY: Apopka PD  
LAST AGENCY INSPECTION: 06/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A1  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 16:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 07/08/2018 19:07:36

---

Diagnostics Check	OK	18:59
Air Blank	0.000	18:59
Control Test	0.079	18:59
Air Blank	0.000	19:00
Subject Sample # 1	0.182	19:03
Breath Volume	1.144	19:03
Air Blank	0.000	19:03
Air Blank	0.000	19:05
Subject Sample # 2	0.189	19:06
Breath Volume	1.394	19:06
Air Blank	0.000	19:06
Control Test	0.078	19:07
Air Blank	0.000	19:07
Diagnostics Check	OK	19:07

---

Data Download Date/Time: 07/11/2018 16:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: HERNANDEZ,YASSELL  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 06/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A1  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 23:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 07/11/2018 00:24:50

---

Diagnostics Check	OK	00:18
Air Blank	0.000	00:19
Control Test	0.080	00:19
Air Blank	0.000	00:19
Subject Sample # 1	0.135	00:20
Breath Volume	1.921	00:20
Air Blank	0.000	00:21
Air Blank	0.000	00:22
Subject Sample # 2	0.131	00:23
Breath Volume	1.554	00:23
Air Blank	0.000	00:23
Control Test	0.080	00:24
Air Blank	0.000	00:24
Diagnostics Check	OK	00:24

---

Data Download Date/Time: 07/11/2018 16:30



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MELVILLE,KELLY,M  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: MOCK,ERIC  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 06/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 08:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:10

SUBJECT'S LAST NAME: KERN

DATE/TIME OF TEST: 07/11/2018 09:48:30

---

Diagnostics Check	OK	09:42
Air Blank	0.000	09:42
Control Test	0.080	09:42
Air Blank	0.000	09:43
Subject Sample # 1	0.000	09:44
Breath Volume	1.210	09:44
Air Blank	0.000	09:44
Air Blank	0.000	09:46
Subject Sample # 2	0.000	09:47
Breath Volume	1.378	09:47
Air Blank	0.000	09:47
Control Test	0.079	09:47
Air Blank	0.000	09:48
Diagnostics Check	OK	09:48

---

Data Download Date/Time: 07/11/2018 16:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MELVILLE,KELLY,M  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: CUSUMANO,ADAM  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 06/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 13:47  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 14:16

SUBJECT'S LAST NAME: CRAIG

DATE/TIME OF TEST: 07/11/2018 14:47:54

---

Diagnostics Check	OK	14:41
Air Blank	0.000	14:42
Control Test	0.079	14:42
Air Blank	0.000	14:43
Subject Sample # 1	0.311	14:43
Breath Volume	1.464	14:43
Air Blank	0.000	14:44
Air Blank	0.000	14:45
Subject Sample # 2	0.310	14:46
Breath Volume	1.964	14:46
Air Blank	0.000	14:46
Control Test	0.079	14:47
Air Blank	0.000	14:47
Diagnostics Check	OK	14:47

---

Data Download Date/Time: 07/11/2018 16:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: DRACH,S  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 06/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 07/12/2018 04:45:42

---

Diagnostics Check	OK	04:39
Air Blank	0.000	04:39
Control Test	0.081	04:40
Air Blank	0.000	04:40
Subject Sample # 1	0.028	04:41
Breath Volume	1.621	04:41
Air Blank	0.000	04:41
Air Blank	0.000	04:43
Subject Sample # 2	0.029	04:44
Breath Volume	2.171	04:44
Air Blank	0.000	04:44
Control Test	0.080	04:45
Air Blank	0.000	04:45
Diagnostics Check	OK	04:45

---

Data Download Date/Time: 07/19/2018 13:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: DRACH,STEFAN  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 06/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 07/13/2018 00:40:52

---

Diagnostics Check		OK	00:34
Air Blank		0.000	00:35
Control Test		0.080	00:35
Air Blank		0.000	00:36
Subject Sample	# 1	0.114	00:36
Breath Volume		2.015	00:36
Air Blank		0.000	00:37
Air Blank		0.000	00:38
Subject Sample	# 2	0.112	00:39
Breath Volume		2.011	00:39
Air Blank		0.000	00:39
Control Test		0.079	00:40
Air Blank		0.000	00:40
Diagnostics Check		OK	00:40

---

Data Download Date/Time: 07/19/2018 13:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLS,ARMAND,G  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: VAUGHN,JACOB  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 06/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 15:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:01

SUBJECT'S LAST NAME: SNYDER

DATE/TIME OF TEST: 07/13/2018 16:37:04

---

Diagnostics Check	OK	16:30
Air Blank	0.000	16:31
Control Test	0.080	16:31
Air Blank	0.000	16:32
Subject Sample # 1	0.000	16:32
Breath Volume	2.566	16:32
Air Blank	0.000	16:33
Air Blank	0.000	16:35
Subject Sample # 2	0.000	16:35
Breath Volume	2.078	16:35
Air Blank	0.000	16:36
Control Test	0.080	16:36
Air Blank	0.000	16:36
Diagnostics Check	OK	16:37

---

Data Download Date/Time: 07/19/2018 13:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMPSON,CALEB,N  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: DRACH,STEFAN  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 06/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 18:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:49

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/13/2018 20:21:15

---

Diagnostics Check		OK	20:12
Air Blank		0.000	20:13
Control Test		0.080	20:13
Air Blank		0.000	20:14
Subject Sample	# 1	0.208	20:16
Breath Volume		2.187	20:16
Air Blank		0.000	20:17
Air Blank		0.000	20:19
Subject Sample	# 2	0.193	20:19
Breath Volume		1.621	20:19
Air Blank		0.000	20:20
Control Test		0.078	20:20
Air Blank		0.000	20:21
Diagnostics Check		OK	20:21

---

Data Download Date/Time: 07/19/2018 13:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CANER,OSVALDO,  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: VENTURA,RONNIE  
ARRESTING OFFICER AGENCY: Edgewood PD  
LAST AGENCY INSPECTION: 06/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: CUNNINGHAM

DATE/TIME OF TEST: 07/14/2018 01:01:14

---

Diagnostics Check		OK	00:53
Air Blank		0.000	00:54
Control Test		0.080	00:54
Air Blank		0.000	00:55
Subject Sample	# 1	0.253	00:56
Breath Volume		1.492	00:56
Air Blank		0.000	00:57
Air Blank		0.000	00:59
Subject Sample	# 2	0.245	00:59
Breath Volume		1.320	00:59
Air Blank		0.000	01:00
Control Test		0.079	01:00
Air Blank		0.000	01:01
Diagnostics Check		OK	01:01

---

Data Download Date/Time: 07/19/2018 13:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: SIMISTER,JAMES  
ARRESTING OFFICER AGENCY: Ocoee PD  
LAST AGENCY INSPECTION: 06/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:55  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: MALONEY

DATE/TIME OF TEST: 07/14/2018 01:26:29

---

Diagnostics Check		OK	01:20
Air Blank		0.000	01:20
Control Test		0.079	01:21
Air Blank		0.000	01:21
Subject Sample	# 1	0.104	01:22
Breath Volume		1.246	01:22
Air Blank		0.000	01:22
Air Blank		0.000	01:24
Subject Sample	# 2	0.100	01:24
Breath Volume		1.347	01:24
Air Blank		0.000	01:25
Control Test		0.078	01:25
Air Blank		0.000	01:26
Diagnostics Check		OK	01:26

---

Data Download Date/Time: 07/19/2018 13:12



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TIAMSON,DAWN,P  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: MOORE,SHELDON  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 06/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: GRAHAM

DATE/TIME OF TEST: 07/14/2018 03:03:58

---

Diagnostics Check		OK	02:49	
Air Blank		0.000	02:49	
Control Test		0.079	02:50	
Air Blank		0.000	02:50	
Subject Sample	# 1	0.130	02:53	Volume Not Met
Breath Volume		0.632	02:53	
Air Blank		0.000	02:54	
Air Blank		0.000	02:56	
Subject Sample	# 2	0.151	02:56	
Breath Volume		1.519	02:56	
Air Blank		0.000	02:57	
Air Blank		0.000	02:59	
Subject Sample	# 3	0.127	03:02	Volume Not Met
Breath Volume		0.738	03:02	
Air Blank		0.000	03:03	
Control Test		0.079	03:03	
Air Blank		0.000	03:03	
Diagnostics Check		OK	03:03	

---

Data Download Date/Time: 07/19/2018 13:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TIAMSON,DAWN,P  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: MOORE,SHELDON  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 06/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: GRAHAM

DATE/TIME OF TEST: 07/14/2018 03:13:58

---

Diagnostics Check		OK	03:05
Air Blank		0.000	03:06
Control Test		0.079	03:06
Air Blank		0.000	03:07
Subject Sample	# 1	0.148	03:07
Breath Volume		1.367	03:07
Air Blank		0.000	03:08
Air Blank		0.000	03:09
Subject Sample	# 2	0.149	03:12
Breath Volume		1.250	03:12
Air Blank		0.000	03:13
Control Test		0.079	03:13
Air Blank		0.000	03:13
Diagnostics Check		OK	03:13

---

Data Download Date/Time: 07/19/2018 13:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: AZEVEDO,FABIO  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 06/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:37  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: BHARATH

DATE/TIME OF TEST: 07/14/2018 05:00:39

---

Diagnostics Check	OK	04:53
Air Blank	0.000	04:53
Control Test	0.080	04:53
Air Blank	0.000	04:54
Subject Sample # 1	0.161	04:55
Breath Volume	1.410	04:55
Air Blank	0.000	04:56
Air Blank	0.000	04:58
Subject Sample # 2	0.164	04:59
Breath Volume	1.820	04:59
Air Blank	0.000	04:59
Control Test	0.078	05:00
Air Blank	0.000	05:00
Diagnostics Check	OK	05:00

---

Data Download Date/Time: 07/19/2018 13:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TIAMSON,DAWN,P  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: COMPTON,LARRY  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 06/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: BUILESVILLA

DATE/TIME OF TEST: 07/15/2018 04:12:53

---

Diagnostics Check	OK	04:05
Air Blank	0.000	04:06
Control Test	0.080	04:06
Air Blank	0.000	04:07
Subject Sample # 1	0.163	04:08
Breath Volume	2.394	04:08
Air Blank	0.000	04:08
Air Blank	0.000	04:10
Subject Sample # 2	0.158	04:11
Breath Volume	1.804	04:11
Air Blank	0.000	04:11
Control Test	0.079	04:12
Air Blank	0.000	04:12
Diagnostics Check	OK	04:12

---

Data Download Date/Time: 07/19/2018 13:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: WEAVER,WILLIAM  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 06/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 04:23  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: MCLEMORE

DATE/TIME OF TEST: 07/15/2018 05:22:51

---

Diagnostics Check	OK	05:11	
Air Blank	0.000	05:11	
Control Test	0.080	05:12	
Air Blank	0.000	05:12	
Subject Sample	# 1	05:15	No Sample Provided
Breath Volume	0.000	05:15	
Air Blank	0.000	05:16	
Air Blank	0.000	05:18	
Subject Sample	# 2	05:21	No Sample Provided
Breath Volume	0.000	05:21	
Air Blank	0.000	05:21	
Control Test	0.080	05:22	
Air Blank	0.000	05:22	
Diagnostics Check	OK	05:22	

---

Data Download Date/Time: 07/19/2018 13:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: WEAVER,WILLIAM  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 06/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 04:23  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: MCLEMORE

DATE/TIME OF TEST: 07/15/2018 05:33:50

---

Diagnostics Check	OK	05:24
Air Blank	0.000	05:24
Control Test	0.079	05:24
Air Blank	0.000	05:25
Subject Sample # 1	0.134	05:26
Breath Volume	1.199	05:26
Air Blank	0.000	05:27
Air Blank	0.000	05:29
Subject Sample # 2	0.133	05:32
Breath Volume	1.554	05:32
Air Blank	0.000	05:32
Control Test	0.079	05:33
Air Blank	0.000	05:33
Diagnostics Check	OK	05:33

---

Data Download Date/Time: 07/19/2018 13:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: WEAVER,WILLIAM  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 06/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: MOLL

DATE/TIME OF TEST: 07/17/2018 01:59:00

---

Diagnostics Check	OK	01:52
Air Blank	0.000	01:53
Control Test	0.080	01:53
Air Blank	0.000	01:54
Subject Sample # 1	0.135	01:54
Breath Volume	1.515	01:54
Air Blank	0.000	01:55
Air Blank	0.000	01:57
Subject Sample # 2	0.132	01:57
Breath Volume	2.574	01:57
Air Blank	0.000	01:58
Control Test	0.079	01:58
Air Blank	0.000	01:58
Diagnostics Check	OK	01:59

---

Data Download Date/Time: 07/19/2018 13:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TIAMSON,DAWN,P  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: JOHNS,MARK  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 06/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 20:10  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:33

SUBJECT'S LAST NAME: HAWKINS

DATE/TIME OF TEST: 07/17/2018 21:03:46

---

Diagnostics Check	OK	20:57
Air Blank	0.000	20:57
Control Test	0.079	20:57
Air Blank	0.000	20:58
Subject Sample # 1	0.000	20:59
Breath Volume	1.105	20:59
Air Blank	0.000	21:00
Air Blank	0.000	21:02
Subject Sample # 2	0.000	21:02
Breath Volume	1.113	21:02
Air Blank	0.000	21:02
Control Test	0.079	21:03
Air Blank	0.000	21:03
Diagnostics Check	OK	21:03

---

Data Download Date/Time: 07/19/2018 13:12



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: OCAMPO,OSCAR  
ARRESTING OFFICER AGENCY: Maitland PD  
LAST AGENCY INSPECTION: 06/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: PERDUE

DATE/TIME OF TEST: 07/19/2018 02:53:32

---

Diagnostics Check	OK	02:47
Air Blank	0.000	02:47
Control Test	0.079	02:48
Air Blank	0.000	02:48
Subject Sample # 1	0.166	02:49
Breath Volume	2.933	02:49
Air Blank	0.000	02:49
Air Blank	0.000	02:51
Subject Sample # 2	0.166	02:52
Breath Volume	2.996	02:52
Air Blank	0.000	02:52
Control Test	0.079	02:52
Air Blank	0.000	02:53
Diagnostics Check	OK	02:53

---

Data Download Date/Time: 07/19/2018 13:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: ELLER,JESSICA  
ARRESTING OFFICER AGENCY: Winter Park PD  
LAST AGENCY INSPECTION: 06/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: MOREHOUSE

DATE/TIME OF TEST: 07/19/2018 04:28:51

---

Diagnostics Check		OK	04:22
Air Blank		0.000	04:22
Control Test		0.079	04:22
Air Blank		0.000	04:23
Subject Sample	# 1	0.206	04:24
Breath Volume		2.097	04:24
Air Blank		0.000	04:25
Air Blank		0.000	04:26
Subject Sample	# 2	0.186	04:27
Breath Volume		1.382	04:27
Air Blank		0.000	04:27
Control Test		0.079	04:28
Air Blank		0.000	04:28
Diagnostics Check		OK	04:28

---

Data Download Date/Time: 07/19/2018 13:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: CORSO,ANTHONY  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 07/19/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: CUNNINGHAM

DATE/TIME OF TEST: 07/21/2018 03:12:06

---

Diagnostics Check	OK	03:05
Air Blank	0.000	03:06
Control Test	0.081	03:06
Air Blank	0.000	03:07
Subject Sample # 1	0.043	03:07
Breath Volume	3.820	03:07
Air Blank	0.000	03:08
Air Blank	0.000	03:10
Subject Sample # 2	0.042	03:10
Breath Volume	3.664	03:10
Air Blank	0.000	03:11
Control Test	0.080	03:11
Air Blank	0.000	03:12
Diagnostics Check	OK	03:12

---

Data Download Date/Time: 07/25/2018 17:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: ANDREWS,GARY  
ARRESTING OFFICER AGENCY: Maitland PD  
LAST AGENCY INSPECTION: 07/19/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: KELLOM

DATE/TIME OF TEST: 07/22/2018 03:14:23

---

Diagnostics Check		OK	03:08
Air Blank		0.000	03:08
Control Test		0.081	03:08
Air Blank		0.000	03:09
Subject Sample	# 1	0.209	03:09
Breath Volume		1.910	03:09
Air Blank		0.000	03:10
Air Blank		0.000	03:12
Subject Sample	# 2	0.210	03:12
Breath Volume		2.527	03:12
Air Blank		0.000	03:13
Control Test		0.080	03:13
Air Blank		0.000	03:14
Diagnostics Check		OK	03:14

---

Data Download Date/Time: 07/25/2018 17:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CANER, OSVALDO,  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: KUZMA, TIMOTHY  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 07/19/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 04:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: HUERTAS RAMIREZ

DATE/TIME OF TEST: 07/23/2018 05:07:32

---

Diagnostics Check	OK	05:01
Air Blank	0.000	05:01
Control Test	0.081	05:02
Air Blank	0.000	05:02
Subject Sample # 1	0.177	05:03
Breath Volume	3.761	05:03
Air Blank	0.000	05:03
Air Blank	0.000	05:05
Subject Sample # 2	0.180	05:06
Breath Volume	3.261	05:06
Air Blank	0.000	05:06
Control Test	0.079	05:06
Air Blank	0.000	05:07
Diagnostics Check	OK	05:07

---

Data Download Date/Time: 07/25/2018 17:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ELBON,NICHOLAS,C  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: ORTIZ,JULIAN  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 07/19/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13917080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: PARRU SOTO

DATE/TIME OF TEST: 07/30/2018 03:38:16

---

Diagnostics Check		OK	03:25	
Air Blank		0.000	03:25	
Control Test		0.080	03:26	
Air Blank		0.000	03:26	
Subject Sample	# 1	0.169	03:29	Volume Not Met
Breath Volume		0.121	03:29	
Air Blank		0.000	03:30	
Air Blank		0.000	03:32	
Subject Sample	# 2	0.177	03:33	
Breath Volume		2.117	03:33	
Air Blank		0.000	03:34	
Air Blank		0.000	03:35	
Subject Sample	# 3	0.184	03:36	
Breath Volume		2.593	03:36	
Air Blank		0.000	03:37	
Control Test		0.079	03:37	
Air Blank		0.000	03:38	
Diagnostics Check		OK	03:38	

---

Data Download Date/Time: 08/14/2018 11:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CASEY,COREY,M  
BREATH TEST OPERATOR AGENCY: Rockledge PD  
ARRESTING OFFICER: GARNER,TYRELL  
ARRESTING OFFICER AGENCY: Rockledge PD  
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 572920  
DRY GAS EXPIRY DATE: 09/13/2019  
VIOLATION CODE: DUI  
ARREST TIME: 12:44  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 13:16

SUBJECT'S LAST NAME: FARRINGTON

DATE/TIME OF TEST: 07/08/2018 13:48:54

---

Diagnostics Check	OK	13:37	
Air Blank	0.000	13:37	
Control Test	0.079	13:38	
Air Blank	0.000	13:38	
Subject Sample	# 1 0.231	13:41	Volume Not Met
Breath Volume	0.035	13:41	
Air Blank	0.000	13:42	
Air Blank	0.000	13:44	
Subject Sample	# 2 0.194	13:47	Volume Not Met
Breath Volume	0.109	13:47	
Air Blank	0.000	13:47	
Control Test	0.077	13:48	
Air Blank	0.000	13:48	
Diagnostics Check	OK	13:48	

---

Data Download Date/Time: 07/31/2018 08:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CASEY,COREY,M  
BREATH TEST OPERATOR AGENCY: Rockledge PD  
ARRESTING OFFICER: GARNER,TYRELL  
ARRESTING OFFICER AGENCY: Rockledge PD  
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 572920  
DRY GAS EXPIRY DATE: 09/13/2019  
VIOLATION CODE: DUI  
ARREST TIME: 12:44  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 13:16

SUBJECT'S LAST NAME: FARRINGTON

DATE/TIME OF TEST: 07/08/2018 13:55:48

---

Diagnostics Check	OK	13:50	
Air Blank	0.000	13:50	
Control Test	0.078	13:51	
Air Blank	0.000	13:51	
Subject Sample	# 1	13:54	Subject Test Refused
Breath Volume	0.000	13:54	
Air Blank	0.000	13:54	
Control Test	0.078	13:55	
Air Blank	0.000	13:55	
Diagnostics Check	OK	13:55	

---

Data Download Date/Time: 07/31/2018 08:18



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLBROOK,NICHOLAS,B  
BREATH TEST OPERATOR AGENCY: Rockledge PD  
ARRESTING OFFICER: LYNCH,ROBERT  
ARRESTING OFFICER AGENCY: Rockledge PD  
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 572920  
DRY GAS EXPIRY DATE: 09/13/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 07/15/2018 02:16:54

---

Diagnostics Check	OK	02:09
Air Blank	0.000	02:10
Control Test	0.080	02:10
Air Blank	0.000	02:11
Subject Sample # 1	0.115	02:12
Breath Volume	1.156	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:14
Subject Sample # 2	0.117	02:15
Breath Volume	1.753	02:15
Air Blank	0.000	02:16
Control Test	0.077	02:16
Air Blank	0.000	02:16
Diagnostics Check	OK	02:16

---

Data Download Date/Time: 07/31/2018 08:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LYNCH,ROBERT,W  
BREATH TEST OPERATOR AGENCY: Rockledge PD  
ARRESTING OFFICER: HOLBROOK,NICHOLAS  
ARRESTING OFFICER AGENCY: Rockledge PD  
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 572920  
DRY GAS EXPIRY DATE: 09/13/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: GATTONI

DATE/TIME OF TEST: 07/20/2018 02:44:27

---

Diagnostics Check	OK	02:37
Air Blank	0.000	02:37
Control Test	0.079	02:38
Air Blank	0.000	02:38
Subject Sample # 1	0.228	02:39
Breath Volume	2.738	02:39
Air Blank	0.000	02:40
Air Blank	0.000	02:42
Subject Sample # 2	0.217	02:42
Breath Volume	2.429	02:42
Air Blank	0.000	02:43
Control Test	0.078	02:43
Air Blank	0.000	02:44
Diagnostics Check	OK	02:44

---

Data Download Date/Time: 07/31/2018 08:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,HECTOR,L  
BREATH TEST OPERATOR AGENCY: Rockledge PD  
ARRESTING OFFICER: GARDNER,TYRELL  
ARRESTING OFFICER AGENCY: Rockledge PD  
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 572920  
DRY GAS EXPIRY DATE: 09/13/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:59  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: CURLEY

DATE/TIME OF TEST: 07/22/2018 03:41:31

---

Diagnostics Check	OK	03:35
Air Blank	0.000	03:35
Control Test	0.079	03:36
Air Blank	0.000	03:36
Subject Sample # 1	0.223	03:37
Breath Volume	2.585	03:37
Air Blank	0.000	03:37
Air Blank	0.000	03:39
Subject Sample # 2	0.212	03:40
Breath Volume	1.914	03:40
Air Blank	0.000	03:40
Control Test	0.078	03:40
Air Blank	0.000	03:41
Diagnostics Check	OK	03:41

---

Data Download Date/Time: 07/31/2018 08:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARLSON,EDWIN,J  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: CROSBY,AUSTIN  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814183  
DRY GAS EXPIRY DATE: 07/26/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 07/01/2018 00:32:24

---

Diagnostics Check	OK	00:25
Air Blank	0.000	00:26
Control Test	0.080	00:26
Air Blank	0.000	00:26
Subject Sample # 1	0.323	00:27
Breath Volume	3.308	00:27
Air Blank	0.000	00:28
Air Blank	0.000	00:29
Subject Sample # 2	0.332	00:30
Breath Volume	2.800	00:30
Air Blank	0.000	00:31
Control Test	0.079	00:31
Air Blank	0.000	00:32
Diagnostics Check	OK	00:32

---

Data Download Date/Time: 07/25/2018 10:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LIVINGSTON,CHRISTOPHER,S  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: LIVINGSTON,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814183  
DRY GAS EXPIRY DATE: 07/26/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: HARTMAN

DATE/TIME OF TEST: 07/03/2018 03:25:11

---

Diagnostics Check		OK	03:18
Air Blank		0.000	03:19
Control Test		0.079	03:19
Air Blank		0.000	03:20
Subject Sample	# 1	0.268	03:20
Breath Volume		1.519	03:20
Air Blank		0.000	03:21
Air Blank		0.000	03:22
Subject Sample	# 2	0.270	03:23
Breath Volume		1.468	03:23
Air Blank		0.000	03:24
Control Test		0.078	03:24
Air Blank		0.000	03:25
Diagnostics Check		OK	03:25

---

Data Download Date/Time: 07/25/2018 10:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LIVINGSTON,CHRISTOPHER,S  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: GUIDONE,PAUL  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814183  
DRY GAS EXPIRY DATE: 07/26/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: RUTOKOWSKI

DATE/TIME OF TEST: 07/04/2018 05:24:44

---

Diagnostics Check	OK	05:18
Air Blank	0.000	05:18
Control Test	0.080	05:18
Air Blank	0.000	05:19
Subject Sample # 1	0.000	05:19
Breath Volume	3.316	05:19
Air Blank	0.000	05:20
Air Blank	0.000	05:22
Subject Sample # 2	0.000	05:23
Breath Volume	2.675	05:23
Air Blank	0.000	05:23
Control Test	0.078	05:24
Air Blank	0.000	05:24
Diagnostics Check	OK	05:24

---

Data Download Date/Time: 07/25/2018 10:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LINEHAN,DANIEL,J  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: LINEHAN,DANIEL  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814183  
DRY GAS EXPIRY DATE: 07/26/2020  
VIOLATION CODE: DUI  
ARREST TIME: 13:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:19

SUBJECT'S LAST NAME: JANKOWSKI

DATE/TIME OF TEST: 07/04/2018 14:56:25

---

Diagnostics Check	OK	14:47
Air Blank	0.000	14:47
Control Test	0.078	14:48
Air Blank	0.000	14:48
Subject Sample # 1	0.253	14:49
Breath Volume	1.925	14:49
Air Blank	0.000	14:50
Air Blank	0.000	14:52
Subject Sample # 2	0.246	14:54
Breath Volume	1.761	14:54
Air Blank	0.000	14:55
Control Test	0.078	14:55
Air Blank	0.000	14:56
Diagnostics Check	OK	14:56

---

Data Download Date/Time: 07/25/2018 10:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARLSON,EDWIN,J  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: CROSBY,AUSTIN  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814183  
DRY GAS EXPIRY DATE: 07/26/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: CURRY

DATE/TIME OF TEST: 07/05/2018 03:04:09

---

Diagnostics Check	OK	02:57
Air Blank	0.000	02:58
Control Test	0.078	02:58
Air Blank	0.000	02:59
Subject Sample # 1	0.167	02:59
Breath Volume	2.699	02:59
Air Blank	0.000	03:00
Air Blank	0.000	03:02
Subject Sample # 2	0.169	03:02
Breath Volume	3.062	03:02
Air Blank	0.000	03:03
Control Test	0.079	03:03
Air Blank	0.000	03:04
Diagnostics Check	OK	03:04

---

Data Download Date/Time: 07/25/2018 10:14



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HESSION,CHRISTOPHER,S  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: HESSION,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814183  
DRY GAS EXPIRY DATE: 07/26/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: FREDRICK

DATE/TIME OF TEST: 07/07/2018 00:13:07

---

Diagnostics Check	OK	00:06
Air Blank	0.000	00:06
Control Test	0.079	00:06
Air Blank	0.000	00:07
Subject Sample # 1	0.122	00:08
Breath Volume	1.496	00:08
Air Blank	0.000	00:08
Air Blank	0.000	00:10
Subject Sample # 2	0.140	00:11
Breath Volume	2.648	00:11
Air Blank	0.000	00:12
Control Test	0.081	00:12
Air Blank	0.000	00:13
Diagnostics Check	OK	00:13

---

Data Download Date/Time: 07/25/2018 10:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LINEHAN,DANIEL,J  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: LINEHAN,DANIEL  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814183  
DRY GAS EXPIRY DATE: 07/26/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 07/07/2018 01:53:10

---

Diagnostics Check	OK	01:46
Air Blank	0.000	01:46
Control Test	0.079	01:46
Air Blank	0.000	01:47
Subject Sample # 1	0.189	01:48
Breath Volume	2.710	01:48
Air Blank	0.000	01:49
Air Blank	0.000	01:50
Subject Sample # 2	0.184	01:51
Breath Volume	1.937	01:51
Air Blank	0.000	01:52
Control Test	0.079	01:52
Air Blank	0.000	01:53
Diagnostics Check	OK	01:53

---

Data Download Date/Time: 07/25/2018 10:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOWARD, WILLIAM,  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: HOWARD, WILLIAM  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814183  
DRY GAS EXPIRY DATE: 07/26/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: GUDMUNDSSON

DATE/TIME OF TEST: 07/07/2018 23:36:56

---

Diagnostics Check	OK	23:30
Air Blank	0.000	23:31
Control Test	0.079	23:31
Air Blank	0.000	23:32
Subject Sample # 1	0.000	23:32
Breath Volume	1.605	23:32
Air Blank	0.000	23:33
Air Blank	0.000	23:35
Subject Sample # 2	0.000	23:35
Breath Volume	1.742	23:35
Air Blank	0.000	23:36
Control Test	0.080	23:36
Air Blank	0.000	23:36
Diagnostics Check	OK	23:36

---

Data Download Date/Time: 07/25/2018 10:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FERGUSON,STEVEN,A  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: DUMMER,LUKE  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814183  
DRY GAS EXPIRY DATE: 07/26/2020  
VIOLATION CODE: DUI  
ARREST TIME: 04:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: CONN

DATE/TIME OF TEST: 07/08/2018 04:53:03

---

Diagnostics Check	OK	04:46
Air Blank	0.000	04:46
Control Test	0.080	04:47
Air Blank	0.000	04:47
Subject Sample # 1	0.000	04:48
Breath Volume	4.265	04:48
Air Blank	0.000	04:49
Air Blank	0.000	04:51
Subject Sample # 2	0.000	04:51
Breath Volume	2.324	04:51
Air Blank	0.000	04:52
Control Test	0.080	04:52
Air Blank	0.000	04:52
Diagnostics Check	OK	04:53

---

Data Download Date/Time: 07/25/2018 10:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LUBRANO,OLIVIA,C  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: RUTHERFORD,CHRIS  
ARRESTING OFFICER AGENCY: West Melbourne PD  
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814183  
DRY GAS EXPIRY DATE: 07/26/2020  
VIOLATION CODE: DUI  
ARREST TIME: 16:48  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 17:25

SUBJECT'S LAST NAME: WARD

DATE/TIME OF TEST: 07/08/2018 17:51:57

---

Diagnostics Check	OK	17:45	
Air Blank	0.000	17:46	
Control Test	0.080	17:46	
Air Blank	0.000	17:47	
Subject Sample # 1	0.312	17:50	Volume Not Met
Breath Volume	0.144	17:50	
Air Blank		17:51	Purge Fail
Air Blank		17:51	Purge Fail

---

Data Download Date/Time: 07/25/2018 10:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LUBRANO,OLIVIA,C  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: RUTHERFORD,CHRIS  
ARRESTING OFFICER AGENCY: West Melbourne PD  
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814183  
DRY GAS EXPIRY DATE: 07/26/2020  
VIOLATION CODE: DUI  
ARREST TIME: 16:48  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 17:25

SUBJECT'S LAST NAME: WARD

DATE/TIME OF TEST: 07/08/2018 18:13:47

---

Diagnostics Check	OK	18:05
Air Blank	0.000	18:05
Control Test	0.078	18:05
Air Blank	0.000	18:06
Subject Sample # 1	0.319	18:09
Breath Volume	2.359	18:09
Air Blank	0.000	18:09
Air Blank	0.000	18:11
Subject Sample # 2	0.335	18:12
Breath Volume	1.710	18:12
Air Blank	0.000	18:12
Control Test	0.079	18:13
Air Blank	0.000	18:13
Diagnostics Check	OK	18:13

---

Data Download Date/Time: 07/25/2018 10:15

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HESSION,CHRISTOPHER,S  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: HESSION,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814183  
DRY GAS EXPIRY DATE: 07/26/2020  
VIOLATION CODE: DUI  
ARREST TIME: 17:19  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 17:44

SUBJECT'S LAST NAME: SIMONETTI

DATE/TIME OF TEST: 07/11/2018 18:13:17

---

Diagnostics Check	OK	18:06
Air Blank	0.000	18:07
Control Test	0.080	18:07
Air Blank	0.000	18:08
Subject Sample # 1	0.187	18:08
Breath Volume	2.472	18:08
Air Blank	0.000	18:09
Air Blank	0.000	18:11
Subject Sample # 2	0.190	18:11
Breath Volume	2.113	18:11
Air Blank	0.000	18:12
Control Test	0.079	18:12
Air Blank	0.000	18:13
Diagnostics Check	OK	18:13

---

Data Download Date/Time: 07/25/2018 10:15

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARLSON,EDWIN,J  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: POOLE,BROOKE  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814183  
DRY GAS EXPIRY DATE: 07/26/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: TULLAR

DATE/TIME OF TEST: 07/19/2018 02:04:39

---

Diagnostics Check	OK	01:57
Air Blank	0.000	01:58
Control Test	0.080	01:58
Air Blank	0.000	01:58
Subject Sample # 1	0.159	02:00
Breath Volume	1.925	02:00
Air Blank	0.000	02:01
Air Blank	0.000	02:02
Subject Sample # 2	0.161	02:03
Breath Volume	1.539	02:03
Air Blank	0.000	02:03
Control Test	0.078	02:04
Air Blank	0.000	02:04
Diagnostics Check	OK	02:04

---

Data Download Date/Time: 07/25/2018 10:15



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LIVINGSTON,CHRISTOPHER,S  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: LIVINGSTON,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814183  
DRY GAS EXPIRY DATE: 07/26/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: FRANCO REZENDE

DATE/TIME OF TEST: 07/21/2018 01:40:04

---

Diagnostics Check		OK	01:28	
Air Blank		0.000	01:29	
Control Test		0.080	01:29	
Air Blank		0.000	01:30	
Subject Sample	# 1	0.190	01:33	Volume Not Met
Breath Volume		0.089	01:33	
Air Blank		0.000	01:34	
Air Blank		0.000	01:35	
Subject Sample	# 2		01:38	Subject Test Refused
Breath Volume		0.027	01:38	
Air Blank		0.000	01:39	
Control Test		0.078	01:39	
Air Blank		0.000	01:39	
Diagnostics Check		OK	01:40	

---

Data Download Date/Time: 07/25/2018 10:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FERGUSON,STEVEN,A  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: TUBAJ,NICHOLAS  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814183  
DRY GAS EXPIRY DATE: 07/26/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: RANDALL

DATE/TIME OF TEST: 07/21/2018 21:07:24

---

Diagnostics Check	OK	21:02	
Air Blank	0.000	21:02	
Control Test	0.079	21:03	
Air Blank	0.000	21:03	
Subject Sample	# 1	21:06	Subject Test Refused
Breath Volume	0.000	21:06	
Air Blank	0.000	21:06	
Control Test	0.080	21:06	
Air Blank	0.000	21:07	
Diagnostics Check	OK	21:07	

---

Data Download Date/Time: 07/25/2018 10:15

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LIVINGSTON,CHRISTOPHER,S  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: DOCKSWELL,TYLER  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814183  
DRY GAS EXPIRY DATE: 07/26/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: PELLICCIO

DATE/TIME OF TEST: 07/23/2018 03:09:50

---

Diagnostics Check		OK	03:03
Air Blank		0.000	03:03
Control Test		0.079	03:04
Air Blank		0.000	03:04
Subject Sample	# 1	0.192	03:05
Breath Volume		3.414	03:05
Air Blank		0.000	03:05
Air Blank		0.000	03:07
Subject Sample	# 2	0.186	03:08
Breath Volume		1.699	03:08
Air Blank		0.000	03:08
Control Test		0.079	03:09
Air Blank		0.000	03:09
Diagnostics Check		OK	03:09

---

Data Download Date/Time: 07/25/2018 10:15

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOWARD, WILLIAM,  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: HOWARD, WILLIAM  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 07/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814183  
DRY GAS EXPIRY DATE: 07/26/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:46

SUBJECT'S LAST NAME: HATHAWAY

DATE/TIME OF TEST: 07/25/2018 22:17:56

---

Diagnostics Check	OK	22:11
Air Blank	0.000	22:12
Control Test	0.080	22:12
Air Blank	0.000	22:12
Subject Sample # 1	0.143	22:13
Breath Volume	1.468	22:13
Air Blank	0.000	22:14
Air Blank	0.000	22:15
Subject Sample # 2	0.141	22:16
Breath Volume	1.625	22:16
Air Blank	0.000	22:17
Control Test	0.080	22:17
Air Blank	0.000	22:17
Diagnostics Check	OK	22:17

---

Data Download Date/Time: 08/23/2018 15:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LIVINGSTON,CHRISTOPHER,S  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: LIVINGSTON,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 07/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814183  
DRY GAS EXPIRY DATE: 07/26/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: GADDY

DATE/TIME OF TEST: 07/30/2018 21:31:27

---

Diagnostics Check	OK	21:25
Air Blank	0.000	21:25
Control Test	0.080	21:25
Air Blank	0.000	21:26
Subject Sample # 1	0.040	21:27
Breath Volume	2.789	21:27
Air Blank	0.000	21:27
Air Blank	0.000	21:29
Subject Sample # 2	0.036	21:29
Breath Volume	2.351	21:29
Air Blank	0.000	21:30
Control Test	0.079	21:30
Air Blank	0.000	21:31
Diagnostics Check	OK	21:31

---

Data Download Date/Time: 08/23/2018 15:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

INSTRUMENT REGISTERED TO: INDIAN HARBOUR PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VERDEROSA, TIFFANY, M  
BREATH TEST OPERATOR AGENCY: Indialantic PD  
ARRESTING OFFICER: VERDEROSA, TIFFANY  
ARRESTING OFFICER AGENCY: Indialantic PD  
LAST AGENCY INSPECTION: 07/15/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03417080A1  
DRY GAS EXPIRY DATE: 03/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 18:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:43

SUBJECT'S LAST NAME: SHAHID

DATE/TIME OF TEST: 07/20/2018 19:39:04

---

Diagnostics Check	OK	19:32
Air Blank	0.000	19:32
Control Test	0.080	19:33
Air Blank	0.000	19:33
Subject Sample # 1	0.145	19:34
Breath Volume	1.847	19:34
Air Blank	0.000	19:35
Air Blank	0.000	19:37
Subject Sample # 2	0.152	19:37
Breath Volume	2.183	19:37
Air Blank	0.000	19:38
Control Test	0.079	19:38
Air Blank	0.000	19:38
Diagnostics Check	OK	19:39

---

Data Download Date/Time: 08/17/2018 04:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

INSTRUMENT REGISTERED TO: INDIAN HARBOUR PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMPSON,ERIC,B  
BREATH TEST OPERATOR AGENCY: Indian Harbour Beach  
ARRESTING OFFICER: GROHAM,JOHN  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 07/15/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03417080A1  
DRY GAS EXPIRY DATE: 03/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 06:35  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 07:04

SUBJECT'S LAST NAME: POWELL

DATE/TIME OF TEST: 07/29/2018 07:31:41

---

Diagnostics Check	OK	07:25
Air Blank	0.000	07:25
Control Test	0.080	07:25
Air Blank	0.000	07:26
Subject Sample # 1	0.007	07:27
Breath Volume	2.050	07:27
Air Blank	0.000	07:27
Air Blank	0.000	07:29
Subject Sample # 2	0.007	07:30
Breath Volume	2.093	07:30
Air Blank	0.000	07:30
Control Test	0.081	07:31
Air Blank	0.000	07:31
Diagnostics Check	OK	07:31

---

Data Download Date/Time: 08/17/2018 04:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES,NICHOLAS,W  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: GRAY,C  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 07/26/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19817080A3  
DRY GAS EXPIRY DATE: 09/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:11  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: GOODSPEED

DATE/TIME OF TEST: 07/28/2018 23:58:55

---

Diagnostics Check		OK	23:43	
Air Blank		0.000	23:43	
Control Test		0.080	23:43	
Air Blank		0.000	23:44	
Subject Sample	# 1	0.217	23:47	Volume Not Met
Breath Volume		0.273	23:47	
Air Blank		0.000	23:48	
Air Blank		0.000	23:49	
Subject Sample	# 2	0.251	23:51	
Breath Volume		1.195	23:51	
Air Blank		0.000	23:52	
Air Blank		0.000	23:53	
Subject Sample	# 3	0.180	23:57	Volume Not Met
Breath Volume		0.378	23:57	
Air Blank		0.000	23:57	
Control Test		0.078	23:58	
Air Blank		0.000	23:58	
Diagnostics Check		OK	23:58	

---

Data Download Date/Time: 08/23/2018 21:22



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES,NICHOLAS,W  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: POOLE,B  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 07/26/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19817080A3  
DRY GAS EXPIRY DATE: 09/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:39

SUBJECT'S LAST NAME: DUGUAY

DATE/TIME OF TEST: 07/29/2018 23:49:59

---

Diagnostics Check	OK	23:42
Air Blank	0.000	23:42
Control Test	0.080	23:43
Air Blank	0.000	23:43
Subject Sample # 1	0.094	23:45
Breath Volume	2.050	23:45
Air Blank	0.000	23:46
Air Blank	0.000	23:48
Subject Sample # 2	0.088	23:48
Breath Volume	2.175	23:48
Air Blank	0.000	23:49
Control Test	0.078	23:49
Air Blank	0.000	23:49
Diagnostics Check	OK	23:49

---

Data Download Date/Time: 08/23/2018 21:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LUTZ,EDWIN,J  
BREATH TEST OPERATOR AGENCY: Palm Bay PD  
ARRESTING OFFICER: LUTZ,EDWIN  
ARRESTING OFFICER AGENCY: Palm Bay PD  
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: BILELLO

DATE/TIME OF TEST: 07/01/2018 21:47:07

---

Diagnostics Check		OK	21:40
Air Blank		0.000	21:41
Control Test		0.081	21:41
Air Blank		0.000	21:41
Subject Sample	# 1	0.234	21:42
Breath Volume		2.753	21:42
Air Blank		0.000	21:43
Air Blank		0.000	21:45
Subject Sample	# 2	0.234	21:45
Breath Volume		2.656	21:45
Air Blank		0.000	21:46
Control Test		0.079	21:46
Air Blank		0.000	21:47
Diagnostics Check		OK	21:47

---

Data Download Date/Time: 07/26/2018 19:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LUTZ,EDWIN,J  
BREATH TEST OPERATOR AGENCY: Palm Bay PD  
ARRESTING OFFICER: LUTZ,EDWIN  
ARRESTING OFFICER AGENCY: Palm Bay PD  
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: GROW

DATE/TIME OF TEST: 07/02/2018 02:42:22

---

Diagnostics Check	OK	02:36
Air Blank	0.000	02:36
Control Test	0.080	02:37
Air Blank	0.000	02:37
Subject Sample # 1	0.199	02:38
Breath Volume	3.300	02:38
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample # 2	0.204	02:40
Breath Volume	3.109	02:40
Air Blank	0.000	02:41
Control Test	0.079	02:41
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

---

Data Download Date/Time: 07/26/2018 19:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LUTZ,EDWIN,J  
BREATH TEST OPERATOR AGENCY: Palm Bay PD  
ARRESTING OFFICER: LUTZ,EDWIN  
ARRESTING OFFICER AGENCY: Palm Bay PD  
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: GAUDY

DATE/TIME OF TEST: 07/05/2018 23:14:40

---

Diagnostics Check	OK	23:08
Air Blank	0.000	23:09
Control Test	0.080	23:09
Air Blank	0.000	23:09
Subject Sample # 1	0.139	23:10
Breath Volume	1.968	23:10
Air Blank	0.000	23:10
Air Blank	0.000	23:12
Subject Sample # 2	0.133	23:13
Breath Volume	1.816	23:13
Air Blank	0.000	23:13
Control Test	0.080	23:14
Air Blank	0.000	23:14
Diagnostics Check	OK	23:14

---

Data Download Date/Time: 07/26/2018 19:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LUTZ,EDWIN,J  
BREATH TEST OPERATOR AGENCY: Palm Bay PD  
ARRESTING OFFICER: LUTZ,EDWIN  
ARRESTING OFFICER AGENCY: Palm Bay PD  
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:33  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: VALERIANI

DATE/TIME OF TEST: 07/07/2018 01:52:03

---

Diagnostics Check	OK	01:46
Air Blank	0.000	01:46
Control Test	0.083	01:46
Air Blank	0.000	01:47
Subject Sample # 1	0.103	01:47
Breath Volume	1.347	01:47
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample # 2	0.101	01:50
Breath Volume	1.585	01:50
Air Blank	0.000	01:51
Control Test	0.080	01:51
Air Blank	0.000	01:51
Diagnostics Check	OK	01:52

---

Data Download Date/Time: 07/26/2018 19:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: BALL,STEPHEN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 22817080A4

DRY GAS EXPIRY DATE: 10/05/2019

VIOLATION CODE: DUI

ARREST TIME: 02:42

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: COLON TORRES

DATE/TIME OF TEST: 07/07/2018 03:12:24

---

Diagnostics Check		OK	03:04
Air Blank		0.000	03:05
Control Test		0.082	03:05
Air Blank		0.000	03:05
Subject Sample	# 1	0.188	03:08
Breath Volume		1.683	03:08
Air Blank		0.000	03:08
Air Blank		0.000	03:10
Subject Sample	# 2	0.190	03:10
Breath Volume		1.726	03:10
Air Blank		0.000	03:11
Control Test		0.081	03:11
Air Blank		0.000	03:12
Diagnostics Check		OK	03:12

---

Data Download Date/Time: 07/26/2018 19:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LUTZ,EDWIN,J  
BREATH TEST OPERATOR AGENCY: Palm Bay PD  
ARRESTING OFFICER: LUTZ,EDWIN  
ARRESTING OFFICER AGENCY: Palm Bay PD  
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 20:32  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:23

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 07/09/2018 20:51:07

---

Diagnostics Check	OK	20:43
Air Blank	0.000	20:44
Control Test	0.080	20:44
Air Blank	0.000	20:45
Subject Sample # 1	0.075	20:46
Breath Volume	2.066	20:46
Air Blank	0.000	20:47
Air Blank	0.000	20:49
Subject Sample # 2	0.075	20:49
Breath Volume	1.535	20:49
Air Blank	0.000	20:50
Control Test	0.078	20:50
Air Blank	0.000	20:51
Diagnostics Check	OK	20:51

---

Data Download Date/Time: 07/26/2018 19:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LUTZ,EDWIN,J  
BREATH TEST OPERATOR AGENCY: Palm Bay PD  
ARRESTING OFFICER: LUTZ,EDWIN  
ARRESTING OFFICER AGENCY: Palm Bay PD  
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: ROLDAN

DATE/TIME OF TEST: 07/09/2018 23:20:31

---

Diagnostics Check	OK	23:14
Air Blank	0.000	23:14
Control Test	0.081	23:14
Air Blank	0.000	23:15
Subject Sample # 1	0.264	23:15
Breath Volume	2.058	23:15
Air Blank	0.000	23:16
Air Blank	0.000	23:18
Subject Sample # 2	0.260	23:18
Breath Volume	1.851	23:18
Air Blank	0.000	23:19
Control Test	0.079	23:19
Air Blank	0.000	23:20
Diagnostics Check	OK	23:20

---

Data Download Date/Time: 07/26/2018 19:42



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LUTZ,EDWIN,J  
BREATH TEST OPERATOR AGENCY: Palm Bay PD  
ARRESTING OFFICER: LUTZ,EDWIN  
ARRESTING OFFICER AGENCY: Palm Bay PD  
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 07/12/2018 23:07:19

---

Diagnostics Check		OK	23:01
Air Blank		0.000	23:01
Control Test		0.083	23:02
Air Blank		0.000	23:02
Subject Sample	# 1	0.184	23:02
Breath Volume		2.597	23:02
Air Blank		0.000	23:03
Air Blank		0.000	23:05
Subject Sample	# 2	0.187	23:05
Breath Volume		1.917	23:05
Air Blank		0.000	23:06
Control Test		0.080	23:06
Air Blank		0.000	23:07
Diagnostics Check		OK	23:07

---

Data Download Date/Time: 07/26/2018 19:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LUTZ,EDWIN,J  
BREATH TEST OPERATOR AGENCY: Palm Bay PD  
ARRESTING OFFICER: LUTZ,EDWIN  
ARRESTING OFFICER AGENCY: Palm Bay PD  
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 20:50  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: PASSLEY

DATE/TIME OF TEST: 07/16/2018 21:05:03

---

Diagnostics Check	OK	20:57
Air Blank	0.000	20:58
Control Test	0.079	20:58
Air Blank	0.000	20:59
Subject Sample # 1	0.091	21:00
Breath Volume	2.082	21:00
Air Blank	0.000	21:01
Air Blank	0.000	21:03
Subject Sample # 2	0.091	21:03
Breath Volume	1.730	21:03
Air Blank	0.000	21:04
Control Test	0.079	21:04
Air Blank	0.000	21:04
Diagnostics Check	OK	21:05

---

Data Download Date/Time: 07/26/2018 19:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LUTZ,EDWIN,J  
BREATH TEST OPERATOR AGENCY: Palm Bay PD  
ARRESTING OFFICER: LUTZ,EDWIN  
ARRESTING OFFICER AGENCY: Palm Bay PD  
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:37  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: REYNOLDS

DATE/TIME OF TEST: 07/24/2018 01:48:50

---

Diagnostics Check	OK	01:41
Air Blank	0.000	01:41
Control Test	0.079	01:42
Air Blank	0.000	01:42
Subject Sample # 1	0.182	01:44
Breath Volume	1.718	01:44
Air Blank	0.000	01:44
Air Blank	0.000	01:46
Subject Sample # 2	0.187	01:47
Breath Volume	1.804	01:47
Air Blank	0.000	01:47
Control Test	0.080	01:48
Air Blank	0.000	01:48
Diagnostics Check	OK	01:48

---

Data Download Date/Time: 07/26/2018 19:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAMILTON,J,W  
BREATH TEST OPERATOR AGENCY: Palm Bay PD  
ARRESTING OFFICER: HAMILTON,J  
ARRESTING OFFICER AGENCY: Palm Bay PD  
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: MACK

DATE/TIME OF TEST: 07/24/2018 02:25:13

---

Diagnostics Check	OK	02:17
Air Blank	0.000	02:17
Control Test	0.081	02:18
Air Blank	0.000	02:18
Subject Sample # 1	0.128	02:19
Breath Volume	1.687	02:19
Air Blank	0.000	02:20
Air Blank	0.000	02:22
Subject Sample # 2	0.122	02:23
Breath Volume	1.437	02:23
Air Blank	0.000	02:24
Control Test	0.079	02:24
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

---

Data Download Date/Time: 07/26/2018 19:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LUTZ,EDWIN,J  
BREATH TEST OPERATOR AGENCY: Palm Bay PD  
ARRESTING OFFICER: LUTZ,EDWIN  
ARRESTING OFFICER AGENCY: Palm Bay PD  
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 07/25/2018 01:12:24

---

Diagnostics Check		OK	01:06
Air Blank		0.000	01:06
Control Test		0.081	01:07
Air Blank		0.000	01:07
Subject Sample	# 1	0.254	01:08
Breath Volume		3.375	01:08
Air Blank		0.000	01:08
Air Blank		0.000	01:10
Subject Sample	# 2	0.255	01:10
Breath Volume		3.382	01:10
Air Blank		0.000	01:11
Control Test		0.078	01:11
Air Blank		0.000	01:12
Diagnostics Check		OK	01:12

---

Data Download Date/Time: 07/26/2018 19:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 22817080A4

DRY GAS EXPIRY DATE: 10/05/2019

VIOLATION CODE: DUI

ARREST TIME: 00:47

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: OKEEFE

DATE/TIME OF TEST: 07/26/2018 01:02:59

---

Diagnostics Check	OK	00:56
Air Blank	0.000	00:57
Control Test	0.080	00:57
Air Blank	0.000	00:58
Subject Sample # 1	0.057	00:58
Breath Volume	3.101	00:58
Air Blank	0.000	00:59
Air Blank	0.000	01:01
Subject Sample # 2	0.052	01:01
Breath Volume	2.710	01:01
Air Blank	0.000	01:02
Control Test	0.077	01:02
Air Blank	0.000	01:02
Diagnostics Check	OK	01:02

---

Data Download Date/Time: 07/26/2018 19:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LUTZ,EDWIN,J  
BREATH TEST OPERATOR AGENCY: Palm Bay PD  
ARRESTING OFFICER: LUTZ,EDWIN  
ARRESTING OFFICER AGENCY: Palm Bay PD  
LAST AGENCY INSPECTION: 07/26/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 20:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:17

SUBJECT'S LAST NAME: VITALIANO

DATE/TIME OF TEST: 07/30/2018 20:43:49

---

Diagnostics Check		OK	20:37
Air Blank		0.000	20:38
Control Test		0.078	20:38
Air Blank		0.000	20:39
Subject Sample	# 1	0.238	20:39
Breath Volume		2.871	20:39
Air Blank		0.000	20:40
Air Blank		0.000	20:41
Subject Sample	# 2	0.235	20:42
Breath Volume		2.437	20:42
Air Blank		0.000	20:42
Control Test		0.080	20:43
Air Blank		0.000	20:43
Diagnostics Check		OK	20:43

---

Data Download Date/Time: 08/28/2018 13:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 07/26/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 22817080A4

DRY GAS EXPIRY DATE: 10/05/2019

VIOLATION CODE: DUI

ARREST TIME: 02:15

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: FAULTERSACK

DATE/TIME OF TEST: 07/31/2018 02:23:47

---

Diagnostics Check	Fail	02:23	Voltage/Current Test
Air Blank	0.000	02:23	

---

Data Download Date/Time: 08/28/2018 13:24



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 07/26/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 22817080A4

DRY GAS EXPIRY DATE: 10/05/2019

VIOLATION CODE: DUI

ARREST TIME: 02:15

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: FAULTERSACK

DATE/TIME OF TEST: 07/31/2018 02:29:05

---

Diagnostics Check	OK	02:25	
Air Blank	0.000	02:25	
Control Test	0.080	02:26	
Air Blank	0.000	02:26	
Subject Sample	# 1	02:27	Subject Test Refused
Breath Volume	0.000	02:27	
Air Blank	0.000	02:28	
Control Test	0.080	02:28	
Air Blank	0.000	02:29	
Diagnostics Check	OK	02:29	

---

Data Download Date/Time: 08/28/2018 13:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KINKOPF,TIMOTHY,E  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: KINKOPF,TIMOTHY  
ARRESTING OFFICER AGENCY: Seminole County SO  
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 18:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:57

SUBJECT'S LAST NAME: MCCOOL

DATE/TIME OF TEST: 07/04/2018 20:23:54

---

Diagnostics Check	OK	20:17
Air Blank	0.000	20:17
Control Test	0.080	20:18
Air Blank	0.000	20:18
Subject Sample # 1	0.125	20:19
Breath Volume	2.023	20:19
Air Blank	0.000	20:19
Air Blank	0.000	20:21
Subject Sample # 2	0.116	20:22
Breath Volume	1.308	20:22
Air Blank	0.000	20:22
Control Test	0.080	20:23
Air Blank	0.000	20:23
Diagnostics Check	OK	20:23

---

Data Download Date/Time: 07/29/2018 21:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARCIA, RAY,  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: RERES, AMANDA  
ARRESTING OFFICER AGENCY: Seminole County SO  
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 20:16  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:47

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 07/04/2018 21:14:39

---

Diagnostics Check	OK	21:08
Air Blank	0.000	21:08
Control Test	0.080	21:08
Air Blank	0.000	21:09
Subject Sample # 1	0.145	21:10
Breath Volume	1.339	21:10
Air Blank	0.000	21:10
Air Blank	0.000	21:12
Subject Sample # 2	0.140	21:13
Breath Volume	1.402	21:13
Air Blank	0.000	21:13
Control Test	0.080	21:14
Air Blank	0.000	21:14
Diagnostics Check	OK	21:14

---

Data Download Date/Time: 07/29/2018 21:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARCIA, RAY,  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: MCCURDY, KATHRYN  
ARRESTING OFFICER AGENCY: Seminole County SO  
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 20:53  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: HOESMAN

DATE/TIME OF TEST: 07/04/2018 22:01:00

---

Diagnostics Check	OK	21:54
Air Blank	0.000	21:55
Control Test	0.081	21:55
Air Blank	0.000	21:56
Subject Sample # 1	0.000	21:56
Breath Volume	1.281	21:56
Air Blank	0.000	21:57
Air Blank	0.000	21:59
Subject Sample # 2	0.000	21:59
Breath Volume	1.375	21:59
Air Blank	0.000	22:00
Control Test	0.081	22:00
Air Blank	0.000	22:00
Diagnostics Check	OK	22:01

---

Data Download Date/Time: 07/29/2018 21:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FERGUSON,RICHARD,A  
BREATH TEST OPERATOR AGENCY: Lake Mary PD  
ARRESTING OFFICER: HALL,JONATHAN  
ARRESTING OFFICER AGENCY: Seminole County SO  
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 19:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:21

SUBJECT'S LAST NAME: TOSOLINI

DATE/TIME OF TEST: 07/08/2018 20:54:36

---

Diagnostics Check	OK	20:47
Air Blank	0.000	20:48
Control Test	0.079	20:48
Air Blank	0.000	20:48
Subject Sample # 1	0.168	20:50
Breath Volume	2.265	20:50
Air Blank	0.000	20:50
Air Blank	0.000	20:52
Subject Sample # 2	0.174	20:53
Breath Volume	1.996	20:53
Air Blank	0.000	20:53
Control Test	0.079	20:54
Air Blank	0.000	20:54
Diagnostics Check	OK	20:54

---

Data Download Date/Time: 07/29/2018 21:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARCIA, RAY,  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: MONANI, MICHAEL  
ARRESTING OFFICER AGENCY: Oviedo PD  
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: NAGY

DATE/TIME OF TEST: 07/09/2018 01:41:35

---

Diagnostics Check	OK	01:35
Air Blank	0.000	01:35
Control Test	0.080	01:36
Air Blank	0.000	01:36
Subject Sample # 1	0.187	01:37
Breath Volume	2.101	01:37
Air Blank	0.000	01:37
Air Blank	0.000	01:39
Subject Sample # 2	0.177	01:40
Breath Volume	2.136	01:40
Air Blank	0.000	01:40
Control Test	0.079	01:41
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

---

Data Download Date/Time: 07/29/2018 21:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ANCION,CALIX,  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: ANCION,CALIX  
ARRESTING OFFICER AGENCY: Seminole County SO  
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 06:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:08

SUBJECT'S LAST NAME: ARISTIZABAL

DATE/TIME OF TEST: 07/10/2018 07:36:46

---

Diagnostics Check	OK	07:30
Air Blank	0.000	07:31
Control Test	0.081	07:31
Air Blank	0.000	07:31
Subject Sample # 1	0.000	07:32
Breath Volume	1.472	07:32
Air Blank	0.000	07:32
Air Blank	0.000	07:34
Subject Sample # 2	0.000	07:35
Breath Volume	1.195	07:35
Air Blank	0.000	07:35
Control Test	0.080	07:36
Air Blank	0.000	07:36
Diagnostics Check	OK	07:36

---

Data Download Date/Time: 07/29/2018 21:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RIGGINS,PAUL,J  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: HILL,SEAN  
ARRESTING OFFICER AGENCY: Sanford PD  
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 17:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 07/16/2018 17:49:59

---

Diagnostics Check	OK	17:43
Air Blank	0.000	17:43
Control Test	0.080	17:43
Air Blank	0.000	17:44
Subject Sample # 1	0.260	17:45
Breath Volume	1.945	17:45
Air Blank	0.000	17:45
Air Blank	0.000	17:47
Subject Sample # 2	0.258	17:48
Breath Volume	1.792	17:48
Air Blank	0.000	17:49
Control Test	0.080	17:49
Air Blank	0.000	17:49
Diagnostics Check	OK	17:49

---

Data Download Date/Time: 07/29/2018 21:29



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RIGGINS,PAUL,J  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: DAIL,EDD  
ARRESTING OFFICER AGENCY: Oviedo PD  
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 08:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:35

SUBJECT'S LAST NAME: PARDO HERNANDEZ

DATE/TIME OF TEST: 07/17/2018 09:10:39

---

Diagnostics Check		OK	08:55	
Air Blank		0.000	08:56	
Control Test		0.080	08:56	
Air Blank		0.000	08:57	
Subject Sample	# 1	0.163	09:00	Volume Not Met
Breath Volume		0.718	09:00	
Air Blank		0.000	09:01	
Air Blank		0.000	09:02	
Subject Sample	# 2	0.175	09:04	
Breath Volume		1.828	09:04	
Air Blank		0.000	09:05	
Air Blank		0.000	09:07	
Subject Sample	# 3	0.165	09:08	
Breath Volume		1.253	09:08	
Air Blank		0.000	09:09	
Control Test		0.078	09:10	
Air Blank		0.000	09:10	
Diagnostics Check		OK	09:10	

---

Data Download Date/Time: 07/29/2018 21:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARCIA, RAY,  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: PRICE, ZACHARY  
ARRESTING OFFICER AGENCY: Casselberry PD  
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:07

SUBJECT'S LAST NAME: ADDISON

DATE/TIME OF TEST: 07/23/2018 05:38:09

---

Diagnostics Check	OK	05:31
Air Blank	0.000	05:32
Control Test	0.081	05:32
Air Blank	0.000	05:33
Subject Sample # 1	0.099	05:33
Breath Volume	2.785	05:33
Air Blank	0.000	05:34
Air Blank	0.000	05:36
Subject Sample # 2	0.098	05:36
Breath Volume	2.617	05:36
Air Blank	0.000	05:37
Control Test	0.080	05:37
Air Blank	0.000	05:38
Diagnostics Check	OK	05:38

---

Data Download Date/Time: 07/29/2018 21:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARCIA, RAY,  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: GUBBA, BRYAN  
ARRESTING OFFICER AGENCY: Seminole County SO  
LAST AGENCY INSPECTION: 07/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: PROB/PAROL  
ARREST TIME: 20:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: WHITAKER

DATE/TIME OF TEST: 07/30/2018 01:11:19

---

Diagnostics Check	OK	01:05
Air Blank	0.000	01:05
Control Test	0.080	01:06
Air Blank	0.000	01:06
Subject Sample # 1	0.103	01:07
Breath Volume	3.031	01:07
Air Blank	0.000	01:07
Air Blank	0.000	01:09
Subject Sample # 2	0.100	01:09
Breath Volume	2.593	01:09
Air Blank	0.000	01:10
Control Test	0.079	01:10
Air Blank	0.000	01:11
Diagnostics Check	OK	01:11

---

Data Download Date/Time: 09/02/2018 20:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARCIA, RAY,  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: KINKOPF, TIM  
ARRESTING OFFICER AGENCY: Seminole County SO  
LAST AGENCY INSPECTION: 07/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: BYRD

DATE/TIME OF TEST: 07/30/2018 04:15:56

---

Diagnostics Check		OK	04:09
Air Blank		0.000	04:09
Control Test		0.081	04:10
Air Blank		0.000	04:10
Subject Sample	# 1	0.215	04:11
Breath Volume		2.593	04:11
Air Blank		0.000	04:11
Air Blank		0.000	04:13
Subject Sample	# 2	0.218	04:14
Breath Volume		2.281	04:14
Air Blank		0.000	04:15
Control Test		0.080	04:15
Air Blank		0.000	04:15
Diagnostics Check		OK	04:15

---

Data Download Date/Time: 09/02/2018 20:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BETHAM,KEITH,D  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: STAPLER,REGINALD  
ARRESTING OFFICER AGENCY: Sanford PD  
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: COLEMAN

DATE/TIME OF TEST: 07/13/2018 01:39:47

---

Diagnostics Check	OK	01:32
Air Blank	0.000	01:32
Control Test	0.082	01:33
Air Blank	0.000	01:33
Subject Sample # 1	0.226	01:34
Breath Volume	1.472	01:34
Air Blank	0.000	01:35
Air Blank	0.000	01:36
Subject Sample # 2	0.226	01:38
Breath Volume	1.191	01:38
Air Blank	0.000	01:38
Control Test	0.081	01:39
Air Blank	0.000	01:39
Diagnostics Check	OK	01:39

---

Data Download Date/Time: 07/29/2018 21:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BETHAM,KEITH,D  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: ROSS,JACOB  
ARRESTING OFFICER AGENCY: Altamonte Springs PD  
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: HAWTER

DATE/TIME OF TEST: 07/13/2018 04:05:21

---

Diagnostics Check	OK	03:58
Air Blank	0.000	03:58
Control Test	0.081	03:59
Air Blank	0.000	03:59
Subject Sample # 1	0.180	04:00
Breath Volume	3.011	04:00
Air Blank	0.000	04:01
Air Blank	0.000	04:03
Subject Sample # 2	0.181	04:03
Breath Volume	2.425	04:03
Air Blank	0.000	04:04
Control Test	0.081	04:04
Air Blank	0.000	04:05
Diagnostics Check	OK	04:05

---

Data Download Date/Time: 07/29/2018 21:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BETHAM,KEITH,D  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: HEIM,EDWARD  
ARRESTING OFFICER AGENCY: Oviedo PD  
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: FIALKOFF

DATE/TIME OF TEST: 07/13/2018 22:24:17

---

Diagnostics Check	OK	22:17
Air Blank	0.000	22:18
Control Test	0.081	22:18
Air Blank	0.000	22:19
Subject Sample # 1	0.105	22:19
Breath Volume	2.875	22:19
Air Blank	0.000	22:20
Air Blank	0.000	22:22
Subject Sample # 2	0.105	22:22
Breath Volume	4.203	22:22
Air Blank	0.000	22:23
Control Test	0.081	22:23
Air Blank	0.000	22:24
Diagnostics Check	OK	22:24

---

Data Download Date/Time: 07/29/2018 21:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: DEROUIN,COLETTE  
ARRESTING OFFICER AGENCY: Altamonte Springs PD  
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 04:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:04

SUBJECT'S LAST NAME: PAUL

DATE/TIME OF TEST: 07/15/2018 05:31:06

---

Diagnostics Check	OK	05:24
Air Blank	0.000	05:25
Control Test	0.081	05:25
Air Blank	0.000	05:25
Subject Sample # 1	0.185	05:26
Breath Volume	1.718	05:26
Air Blank	0.000	05:27
Air Blank	0.000	05:29
Subject Sample # 2	0.181	05:29
Breath Volume	2.906	05:29
Air Blank	0.000	05:30
Control Test	0.081	05:30
Air Blank	0.000	05:31
Diagnostics Check	OK	05:31

---

Data Download Date/Time: 07/29/2018 21:32



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REHDER,PATRICK,W  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: HOLLAND,TROY  
ARRESTING OFFICER AGENCY: Seminole County SO  
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 19:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:08

SUBJECT'S LAST NAME: MILLARD

DATE/TIME OF TEST: 07/20/2018 20:32:52

---

Diagnostics Check	OK	20:30	
Air Blank	0.000	20:31	
Control Test	0.081	20:31	
Air Blank	0.000	20:32	
Subject Sample	# 1	20:32	Sequence Aborted
Breath Volume	0.000	20:32	
Air Blank	0.000	20:32	

---

Data Download Date/Time: 07/29/2018 21:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: WARNER,MICHAEL  
ARRESTING OFFICER AGENCY: Maitland PD  
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: LAWLOR

DATE/TIME OF TEST: 07/21/2018 04:26:19

---

Diagnostics Check	OK	04:20
Air Blank	0.000	04:20
Control Test	0.081	04:20
Air Blank	0.000	04:21
Subject Sample # 1	0.115	04:21
Breath Volume	1.406	04:21
Air Blank	0.000	04:22
Air Blank	0.000	04:24
Subject Sample # 2	0.122	04:24
Breath Volume	1.910	04:24
Air Blank	0.000	04:25
Control Test	0.081	04:25
Air Blank	0.000	04:26
Diagnostics Check	OK	04:26

---

Data Download Date/Time: 07/29/2018 21:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: HOLLAND,TROY  
ARRESTING OFFICER AGENCY: Seminole County SO  
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: CORRALES

DATE/TIME OF TEST: 07/22/2018 01:17:30

---

Diagnostics Check	OK	01:11
Air Blank	0.000	01:11
Control Test	0.082	01:12
Air Blank	0.000	01:12
Subject Sample # 1	0.094	01:13
Breath Volume	1.574	01:13
Air Blank	0.000	01:13
Air Blank	0.000	01:15
Subject Sample # 2	0.100	01:15
Breath Volume	4.101	01:15
Air Blank	0.000	01:16
Control Test	0.082	01:16
Air Blank	0.000	01:17
Diagnostics Check	OK	01:17

---

Data Download Date/Time: 07/29/2018 21:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARCIA, RAY,  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: BROWN, JOEL  
ARRESTING OFFICER AGENCY: Oviedo PD  
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:59  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: MCMANAMA

DATE/TIME OF TEST: 07/25/2018 01:03:18

---

Diagnostics Check	OK	00:56
Air Blank	0.000	00:56
Control Test	0.080	00:56
Air Blank	0.000	00:57
Subject Sample # 1	0.171	00:58
Breath Volume	2.566	00:58
Air Blank	0.000	00:58
Air Blank	0.000	01:00
Subject Sample # 2	0.166	01:01
Breath Volume	1.191	01:01
Air Blank	0.000	01:02
Control Test	0.081	01:02
Air Blank	0.000	01:03
Diagnostics Check	OK	01:03

---

Data Download Date/Time: 07/29/2018 21:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARCIA, RAY,  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: LIEBKNECHT, KEVIN  
ARRESTING OFFICER AGENCY: Maitland PD  
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: CONKLIN

DATE/TIME OF TEST: 07/25/2018 03:37:17

---

Diagnostics Check		OK	03:30
Air Blank		0.000	03:31
Control Test		0.082	03:31
Air Blank		0.000	03:31
Subject Sample	# 1	0.106	03:32
Breath Volume		1.773	03:32
Air Blank		0.000	03:33
Air Blank		0.000	03:35
Subject Sample	# 2	0.110	03:35
Breath Volume		2.796	03:35
Air Blank		0.000	03:36
Control Test		0.082	03:36
Air Blank		0.000	03:37
Diagnostics Check		OK	03:37

---

Data Download Date/Time: 07/29/2018 21:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BETHAM,KEITH,D  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: MEYERS,MARY  
ARRESTING OFFICER AGENCY: Seminole County SO  
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:13  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: CIABATTONI

DATE/TIME OF TEST: 07/28/2018 03:27:15

---

Diagnostics Check	OK	03:19
Air Blank	0.000	03:19
Control Test	0.081	03:20
Air Blank	0.000	03:20
Subject Sample # 1	0.169	03:21
Breath Volume	1.812	03:21
Air Blank	0.000	03:22
Air Blank	0.000	03:24
Subject Sample # 2	0.166	03:25
Breath Volume	1.589	03:25
Air Blank	0.000	03:26
Control Test	0.081	03:26
Air Blank	0.000	03:27
Diagnostics Check	OK	03:27

---

Data Download Date/Time: 07/29/2018 21:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KASELAK,STEPHEN,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: KASELAK,S  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG815602  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: BURAK

DATE/TIME OF TEST: 07/04/2018 01:05:40

---

Diagnostics Check	OK	01:02	
Air Blank	0.000	01:02	
Control Test	0.079	01:03	
Air Blank	0.000	01:03	
Subject Sample	# 1	01:04	Subject Test Refused
Breath Volume	0.000	01:04	
Air Blank	0.000	01:04	
Control Test	0.079	01:05	
Air Blank	0.000	01:05	
Diagnostics Check	OK	01:05	

---

Data Download Date/Time: 07/25/2018 21:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ELLIS,A,M  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: ELLIS,A  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG815602  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: PICKFORD

DATE/TIME OF TEST: 07/04/2018 02:44:21

---

Diagnostics Check	OK	02:40	
Air Blank	0.000	02:40	
Control Test	0.078	02:40	
Air Blank	0.000	02:41	
Subject Sample	# 1	02:42	Subject Test Refused
Breath Volume	0.000	02:42	
Air Blank	0.000	02:43	
Control Test	0.078	02:43	
Air Blank	0.000	02:44	
Diagnostics Check	OK	02:44	

---

Data Download Date/Time: 07/25/2018 21:23