

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: COFFMAN,KYLE
ARRESTING OFFICER AGENCY: Oviedo PD
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: BONHAM

DATE/TIME OF TEST: 07/04/2018 00:32:58

Diagnostics Check	OK	00:25
Air Blank	0.000	00:26
Control Test	0.079	00:26
Air Blank	0.000	00:27
Subject Sample # 1	0.152	00:27
Breath Volume	1.117	00:27
Air Blank	0.000	00:28
Air Blank	0.000	00:30
Subject Sample # 2	0.151	00:31
Breath Volume	1.289	00:31
Air Blank	0.000	00:32
Control Test	0.079	00:32
Air Blank	0.000	00:32
Diagnostics Check	OK	00:32

Data Download Date/Time: 07/29/2018 21:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: HALL,JONATHAN
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: HERNANDEZ DELGADO

DATE/TIME OF TEST: 07/04/2018 04:06:21

Diagnostics Check	OK	03:57
Air Blank	0.000	03:58
Control Test	0.079	03:58
Air Blank	0.000	03:59
Subject Sample # 1	0.138	04:01
Breath Volume	1.363	04:01
Air Blank	0.000	04:02
Air Blank	0.000	04:04
Subject Sample # 2	0.145	04:04
Breath Volume	1.234	04:04
Air Blank	0.000	04:05
Control Test	0.078	04:05
Air Blank	0.000	04:06
Diagnostics Check	OK	04:06

Data Download Date/Time: 07/29/2018 21:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: ROSS,JACOB
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: GOGLAS

DATE/TIME OF TEST: 07/12/2018 02:27:21

Diagnostics Check	OK	02:19
Air Blank	0.000	02:20
Control Test	0.078	02:20
Air Blank	0.000	02:21
Subject Sample # 1	0.327	02:22
Breath Volume	1.867	02:22
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample # 2	0.324	02:25
Breath Volume	2.695	02:25
Air Blank	0.000	02:26
Control Test	0.078	02:26
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

Data Download Date/Time: 07/29/2018 21:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINKOPF,TIMOTHY,E
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: KINKOPF,TIMOTHY
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 16:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: BRINKER

DATE/TIME OF TEST: 07/19/2018 19:59:05

Diagnostics Check	OK	19:46	
Air Blank	0.000	19:46	
Control Test	0.078	19:46	
Air Blank	0.000	19:47	
Subject Sample	# 1 0.213	19:48	
Breath Volume	2.511	19:48	
Air Blank	0.000	19:49	
Air Blank	0.000	19:50	
Subject Sample	# 2 0.166	19:54	Volume Not Met
Breath Volume	0.375	19:54	
Air Blank	0.000	19:54	
Air Blank	0.000	19:56	
Subject Sample	# 3 0.208	19:57	
Breath Volume	1.593	19:57	
Air Blank	0.000	19:58	
Control Test	0.080	19:58	
Air Blank	0.000	19:58	
Diagnostics Check	OK	19:59	

Data Download Date/Time: 07/29/2018 21:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: WARNER,MICHAEL
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: MCCLLOUD

DATE/TIME OF TEST: 07/20/2018 04:38:17

Diagnostics Check	OK	04:32
Air Blank	0.000	04:32
Control Test	0.080	04:32
Air Blank	0.000	04:33
Subject Sample # 1	0.180	04:33
Breath Volume	1.343	04:33
Air Blank	0.000	04:34
Air Blank	0.000	04:36
Subject Sample # 2	0.180	04:36
Breath Volume	1.320	04:36
Air Blank	0.000	04:37
Control Test	0.080	04:37
Air Blank	0.000	04:38
Diagnostics Check	OK	04:38

Data Download Date/Time: 07/29/2018 21:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001000

INSTRUMENT REGISTERED TO: COCOA P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HALL,NICHOLAS,A

BREATH TEST OPERATOR AGENCY: Cocoa PD

ARRESTING OFFICER: RASEY,DANIEL

ARRESTING OFFICER AGENCY: FFWCC (Ocala)

LAST AGENCY INSPECTION: 06/26/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 02618080A3

DRY GAS EXPIRY DATE: 03/05/2020

VIOLATION CODE: DUI

ARREST TIME: 18:59

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:46

SUBJECT'S LAST NAME: SHAW

DATE/TIME OF TEST: 07/01/2018 20:15:07

Diagnostics Check	OK	20:09
Air Blank	0.000	20:09
Control Test	0.078	20:10
Air Blank	0.000	20:10
Subject Sample # 1	0.000	20:11
Breath Volume	2.285	20:11
Air Blank	0.000	20:11
Air Blank	0.000	20:13
Subject Sample # 2	0.000	20:13
Breath Volume	2.027	20:13
Air Blank	0.000	20:14
Control Test	0.078	20:14
Air Blank	0.000	20:15
Diagnostics Check	OK	20:15

Data Download Date/Time: 07/27/2018 20:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001000

INSTRUMENT REGISTERED TO: COCOA P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HALL,NICHOLAS,A

BREATH TEST OPERATOR AGENCY: Cocoa PD

ARRESTING OFFICER: HALL,NICHOLAS

ARRESTING OFFICER AGENCY: Cocoa PD

LAST AGENCY INSPECTION: 06/26/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 02618080A3

DRY GAS EXPIRY DATE: 03/05/2020

VIOLATION CODE: DUI

ARREST TIME: 21:56

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: AYALA

DATE/TIME OF TEST: 07/01/2018 22:16:07

Diagnostics Check	OK	22:12	
Air Blank	0.000	22:13	
Control Test	0.079	22:13	
Air Blank	0.000	22:14	
Subject Sample	# 1	22:14	Subject Test Refused
Breath Volume	0.000	22:14	
Air Blank	0.000	22:15	
Control Test	0.079	22:15	
Air Blank	0.000	22:16	
Diagnostics Check	OK	22:16	

Data Download Date/Time: 07/27/2018 20:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001000

INSTRUMENT REGISTERED TO: COCOA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PICHECO,JOHN,A
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: BRADSHAW,CHRISTOPHER
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 06/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: DEGATINA

DATE/TIME OF TEST: 07/04/2018 02:56:49

Diagnostics Check	OK	02:45	
Air Blank	0.000	02:45	
Control Test	0.079	02:45	
Air Blank	0.000	02:46	
Subject Sample	# 1 0.204	02:49	Volume Not Met
Breath Volume	0.121	02:49	
Air Blank	0.000	02:50	
Air Blank	0.000	02:51	
Subject Sample	# 2 0.234	02:52	
Breath Volume	1.703	02:52	
Air Blank	0.000	02:53	
Air Blank	0.000	02:54	
Subject Sample	# 3 0.237	02:55	
Breath Volume	2.371	02:55	
Air Blank	0.000	02:55	
Control Test	0.074	02:56	Control Outside Tolerance
Air Blank	0.000	02:56	

Data Download Date/Time: 07/27/2018 20:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001000

INSTRUMENT REGISTERED TO: COCOA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINSTEAD,JUSTIN,M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: HERNANDEZ,LUIS
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 06/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: JUAN

DATE/TIME OF TEST: 07/04/2018 22:45:22

Diagnostics Check	OK	22:36
Air Blank	0.000	22:36
Control Test	0.079	22:36
Air Blank	0.000	22:37
Subject Sample # 1	0.253	22:39
Breath Volume	1.875	22:39
Air Blank	0.000	22:40
Air Blank	0.000	22:42
Subject Sample # 2	0.258	22:43
Breath Volume	2.546	22:43
Air Blank	0.000	22:44
Control Test	0.077	22:44
Air Blank	0.000	22:45
Diagnostics Check	OK	22:45

Data Download Date/Time: 07/27/2018 20:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001000

INSTRUMENT REGISTERED TO: COCOA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,JAMES,H
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: HERNANDEZ,L
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 06/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: CURTIN

DATE/TIME OF TEST: 07/06/2018 23:43:26

Diagnostics Check	OK	23:40	
Air Blank	0.000	23:41	
Control Test	0.080	23:41	
Air Blank	0.000	23:41	
Subject Sample	# 1	23:42	Subject Test Refused
Breath Volume	0.000	23:42	
Air Blank	0.000	23:42	
Control Test	0.079	23:42	
Air Blank	0.000	23:43	
Diagnostics Check	OK	23:43	

Data Download Date/Time: 07/27/2018 20:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001000

INSTRUMENT REGISTERED TO: COCOA P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HALL,NICHOLAS,A

BREATH TEST OPERATOR AGENCY: Cocoa PD

ARRESTING OFFICER: WADE,GLENN

ARRESTING OFFICER AGENCY: Cocoa PD

LAST AGENCY INSPECTION: 06/26/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 02618080A3

DRY GAS EXPIRY DATE: 03/05/2020

VIOLATION CODE: DUI

ARREST TIME: 20:10

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:27

SUBJECT'S LAST NAME: RENFROE

DATE/TIME OF TEST: 07/24/2018 20:53:06

Diagnostics Check	OK	20:48	
Air Blank	0.000	20:48	
Control Test	0.078	20:49	
Air Blank	0.000	20:49	
Subject Sample	# 1	20:51	Subject Test Refused
Breath Volume	0.019	20:51	
Air Blank	0.000	20:52	
Control Test	0.079	20:52	
Air Blank	0.000	20:53	
Diagnostics Check	OK	20:53	

Data Download Date/Time: 07/27/2018 20:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD, WILLIAM,
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: HOWARD, WILLIAM
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: MESTAS

DATE/TIME OF TEST: 07/08/2018 03:54:21

Diagnostics Check	OK	03:47
Air Blank	0.000	03:47
Control Test	0.078	03:48
Air Blank	0.000	03:48
Subject Sample # 1	0.087	03:49
Breath Volume	1.394	03:49
Air Blank	0.000	03:50
Air Blank	0.000	03:51
Subject Sample # 2	0.092	03:52
Breath Volume	2.250	03:52
Air Blank	0.000	03:53
Control Test	0.078	03:53
Air Blank	0.000	03:54
Diagnostics Check	OK	03:54

Data Download Date/Time: 07/28/2018 05:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCRANN,SEAN,M
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: MCCRANN,SEAN
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: SPINELLI

DATE/TIME OF TEST: 07/14/2018 02:43:45

Diagnostics Check	OK	02:37
Air Blank	0.000	02:37
Control Test	0.077	02:38
Air Blank	0.000	02:38
Subject Sample # 1	0.140	02:39
Breath Volume	1.925	02:39
Air Blank	0.000	02:39
Air Blank	0.000	02:41
Subject Sample # 2	0.131	02:42
Breath Volume	1.867	02:42
Air Blank	0.000	02:42
Control Test	0.077	02:43
Air Blank	0.000	02:43
Diagnostics Check	OK	02:43

Data Download Date/Time: 07/28/2018 05:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD, WILLIAM,
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: HOWARD, WILLIAM
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: SWEENEY

DATE/TIME OF TEST: 07/14/2018 03:45:01

Diagnostics Check	OK	03:38
Air Blank	0.000	03:38
Control Test	0.077	03:39
Air Blank	0.000	03:39
Subject Sample # 1	0.065	03:40
Breath Volume	1.953	03:40
Air Blank	0.000	03:41
Air Blank	0.000	03:43
Subject Sample # 2	0.067	03:43
Breath Volume	1.746	03:43
Air Blank	0.000	03:44
Control Test	0.078	03:44
Air Blank	0.000	03:44
Diagnostics Check	OK	03:45

Data Download Date/Time: 07/28/2018 05:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, ANTHONY, M
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: HOLLAND, ANTHONY
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 09:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:20

SUBJECT'S LAST NAME: WORTHLEY

DATE/TIME OF TEST: 07/15/2018 10:50:07

Diagnostics Check	OK	10:42
Air Blank	0.000	10:43
Control Test	0.077	10:43
Air Blank	0.000	10:43
Subject Sample # 1	0.200	10:44
Breath Volume	2.199	10:44
Air Blank	0.000	10:45
Air Blank	0.000	10:47
Subject Sample # 2	0.186	10:48
Breath Volume	1.515	10:48
Air Blank	0.000	10:49
Control Test	0.077	10:49
Air Blank	0.000	10:50
Diagnostics Check	OK	10:50

Data Download Date/Time: 07/28/2018 05:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOUSMAN,JOHN,D
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: COCCHIOLA,AUSTIN
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 15:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: CALDWELL

DATE/TIME OF TEST: 07/17/2018 16:52:18

Diagnostics Check	OK	16:43	
Air Blank	0.000	16:44	
Control Test	0.076	16:44	
Air Blank	0.000	16:45	
Subject Sample	# 1 0.240	16:48	Volume Not Met
Breath Volume	0.257	16:48	
Air Blank	0.000	16:48	
Air Blank	0.000	16:50	
Subject Sample	# 2	16:50	Subject Test Refused
Breath Volume	0.000	16:50	
Air Blank	0.000	16:51	
Control Test	0.074	16:51	Control Outside Tolerance
Air Blank	0.000	16:52	

Data Download Date/Time: 07/28/2018 05:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIEBELL,DON,R
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: TRIEBELL,DON
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: JULIEN

DATE/TIME OF TEST: 07/21/2018 23:12:18

Diagnostics Check	OK	23:04
Air Blank	0.000	23:04
Control Test	0.077	23:05
Air Blank	0.000	23:05
Subject Sample # 1	0.152	23:07
Breath Volume	2.175	23:07
Air Blank	0.000	23:07
Air Blank	0.000	23:09
Subject Sample # 2	0.160	23:10
Breath Volume	3.527	23:10
Air Blank	0.000	23:11
Control Test	0.076	23:11
Air Blank	0.000	23:12
Diagnostics Check	OK	23:12

Data Download Date/Time: 07/28/2018 05:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER,JOHN,T
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: SPENCER,ROBERT
ARRESTING OFFICER AGENCY: Wauchula PD
LAST AGENCY INSPECTION: 06/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: OTHER
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: JIMINEZ

DATE/TIME OF TEST: 07/01/2018 00:30:50

Diagnostics Check	OK	00:25	
Air Blank	0.000	00:26	
Control Test	0.080	00:26	
Air Blank	0.000	00:27	
Subject Sample	# 1	00:29	Subject Test Refused
Breath Volume	0.023	00:29	
Air Blank	0.000	00:29	
Control Test	0.081	00:30	
Air Blank	0.000	00:30	
Diagnostics Check	OK	00:30	

Data Download Date/Time: 07/29/2018 13:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTWRIGHT JR,GARY,W
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: JOHNSON,CIERRA
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 06/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: OTHER
ARREST TIME: 22:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: MELTON

DATE/TIME OF TEST: 07/06/2018 23:40:13

Diagnostics Check	OK	23:34
Air Blank	0.000	23:34
Control Test	0.081	23:34
Air Blank	0.000	23:35
Subject Sample # 1	0.239	23:35
Breath Volume	1.578	23:35
Air Blank	0.000	23:36
Air Blank	0.000	23:38
Subject Sample # 2	0.236	23:38
Breath Volume	1.316	23:38
Air Blank	0.000	23:39
Control Test	0.081	23:39
Air Blank	0.000	23:40
Diagnostics Check	OK	23:40

Data Download Date/Time: 07/29/2018 13:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: STEVENSON

DATE/TIME OF TEST: 07/01/2018 02:11:48

Diagnostics Check	OK	02:05
Air Blank	0.000	02:05
Control Test	0.077	02:06
Air Blank	0.000	02:06
Subject Sample # 1	0.188	02:07
Breath Volume	2.652	02:07
Air Blank	0.000	02:07
Air Blank	0.000	02:09
Subject Sample # 2	0.190	02:10
Breath Volume	2.027	02:10
Air Blank	0.000	02:10
Control Test	0.079	02:11
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

Data Download Date/Time: 07/25/2018 23:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, WILLIAM, T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: BERGER

DATE/TIME OF TEST: 07/01/2018 04:02:14

Diagnostics Check	OK	03:55
Air Blank	0.000	03:56
Control Test	0.080	03:56
Air Blank	0.000	03:57
Subject Sample # 1	0.093	03:57
Breath Volume	4.492	03:57
Air Blank	0.000	03:58
Air Blank	0.000	04:00
Subject Sample # 2	0.093	04:00
Breath Volume	4.246	04:00
Air Blank	0.000	04:01
Control Test	0.080	04:01
Air Blank	0.000	04:02
Diagnostics Check	OK	04:02

Data Download Date/Time: 07/25/2018 23:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: SIEV

DATE/TIME OF TEST: 07/04/2018 02:51:14

Diagnostics Check	OK	02:45
Air Blank	0.000	02:45
Control Test	0.079	02:45
Air Blank	0.000	02:46
Subject Sample # 1	0.208	02:46
Breath Volume	2.539	02:46
Air Blank	0.000	02:47
Air Blank	0.000	02:49
Subject Sample # 2	0.200	02:49
Breath Volume	2.429	02:49
Air Blank	0.000	02:50
Control Test	0.079	02:50
Air Blank	0.000	02:51
Diagnostics Check	OK	02:51

Data Download Date/Time: 07/25/2018 23:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNAVELY,KENT,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SNAVELY,KENT
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 10:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:45

SUBJECT'S LAST NAME: WALLEN

DATE/TIME OF TEST: 07/04/2018 11:12:23

Diagnostics Check	OK	11:06
Air Blank	0.000	11:06
Control Test	0.080	11:06
Air Blank	0.000	11:07
Subject Sample # 1	0.000	11:08
Breath Volume	2.148	11:08
Air Blank	0.000	11:08
Air Blank	0.000	11:10
Subject Sample # 2	0.000	11:11
Breath Volume	1.757	11:11
Air Blank	0.000	11:11
Control Test	0.079	11:11
Air Blank	0.000	11:12
Diagnostics Check	OK	11:12

Data Download Date/Time: 07/25/2018 23:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: COCHRAN

DATE/TIME OF TEST: 07/05/2018 00:01:54

Diagnostics Check	OK	23:55
Air Blank	0.000	23:55
Control Test	0.080	23:55
Air Blank	0.000	23:56
Subject Sample # 2	0.125	23:57
Breath Volume	3.289	23:57
Air Blank	0.000	23:58
Air Blank	0.000	23:59
Subject Sample # 1	0.120	00:00
Breath Volume	2.703	00:00
Air Blank	0.000	00:01
Control Test	0.079	00:01
Air Blank	0.000	00:01
Diagnostics Check	OK	00:01

Data Download Date/Time: 07/25/2018 23:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, WILLIAM, T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 07/05/2018 02:29:47

Diagnostics Check	OK	02:22
Air Blank	0.000	02:23
Control Test	0.079	02:23
Air Blank	0.000	02:23
Subject Sample # 1	0.141	02:25
Breath Volume	2.914	02:25
Air Blank	0.000	02:25
Air Blank	0.000	02:27
Subject Sample # 2	0.136	02:28
Breath Volume	2.839	02:28
Air Blank	0.000	02:28
Control Test	0.080	02:29
Air Blank	0.000	02:29
Diagnostics Check	OK	02:29

Data Download Date/Time: 07/25/2018 23:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE,DEPUTY WILLIAM,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WILTSE,DEPUTY WILLIAM
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: ABRAHAMIAN

DATE/TIME OF TEST: 07/05/2018 04:02:54

Diagnostics Check	OK	04:00	
Air Blank	0.000	04:00	
Control Test	0.079	04:00	
Air Blank	0.000	04:01	
Subject Sample	# 1	04:01	Subject Test Refused
Breath Volume	0.000	04:01	
Air Blank	0.000	04:02	
Control Test	0.079	04:02	
Air Blank	0.000	04:02	
Diagnostics Check	OK	04:02	

Data Download Date/Time: 07/25/2018 23:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: SHUFORD

DATE/TIME OF TEST: 07/06/2018 00:31:42

Diagnostics Check	OK	00:25
Air Blank	0.000	00:25
Control Test	0.080	00:26
Air Blank	0.000	00:26
Subject Sample # 1	0.065	00:27
Breath Volume	1.699	00:27
Air Blank	0.000	00:27
Air Blank	0.000	00:29
Subject Sample # 2	0.067	00:30
Breath Volume	1.515	00:30
Air Blank	0.000	00:30
Control Test	0.079	00:31
Air Blank	0.000	00:31
Diagnostics Check	OK	00:31

Data Download Date/Time: 07/25/2018 23:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 07/06/2018 04:13:56

Diagnostics Check	OK	04:07
Air Blank	0.000	04:08
Control Test	0.080	04:08
Air Blank	0.000	04:09
Subject Sample # 1	0.165	04:09
Breath Volume	2.984	04:09
Air Blank	0.000	04:10
Air Blank	0.000	04:12
Subject Sample # 2	0.158	04:12
Breath Volume	2.187	04:12
Air Blank	0.000	04:13
Control Test	0.079	04:13
Air Blank	0.000	04:13
Diagnostics Check	OK	04:13

Data Download Date/Time: 07/25/2018 23:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: HOLLANDER

DATE/TIME OF TEST: 07/07/2018 00:58:58

Diagnostics Check	OK	00:46	
Air Blank	0.000	00:46	
Control Test	0.080	00:47	
Air Blank	0.000	00:47	
Subject Sample	# 1	00:50	No Sample Provided
Breath Volume	0.000	00:50	
Air Blank	0.000	00:51	
Air Blank	0.000	00:53	
Subject Sample	# 2	0.125	00:53
Breath Volume	2.015	00:53	
Air Blank	0.000	00:54	
Air Blank	0.000	00:56	
Subject Sample	# 3	0.121	00:57
Breath Volume	2.656	00:57	
Air Blank	0.000	00:58	
Control Test	0.080	00:58	
Air Blank	0.000	00:58	
Diagnostics Check	OK	00:58	

Data Download Date/Time: 07/25/2018 23:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 07/08/2018 01:30:39

Diagnostics Check	OK	01:27	
Air Blank	0.000	01:28	
Control Test	0.079	01:28	
Air Blank	0.000	01:29	
Subject Sample	# 1	01:29	Subject Test Refused
Breath Volume	0.000	01:29	
Air Blank	0.000	01:29	
Control Test	0.079	01:30	
Air Blank	0.000	01:30	
Diagnostics Check	OK	01:30	

Data Download Date/Time: 07/25/2018 23:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 05:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:07

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 07/08/2018 05:46:15

Diagnostics Check	OK	05:34	
Air Blank	0.000	05:35	
Control Test	0.080	05:35	
Air Blank	0.000	05:35	
Subject Sample	# 1 0.182	05:38	Volume Not Met
Breath Volume	0.191	05:38	
Air Blank	0.000	05:39	
Air Blank	0.000	05:41	
Subject Sample	# 2 0.184	05:44	Volume Not Met
Breath Volume	0.394	05:44	
Air Blank	0.000	05:45	
Control Test	0.079	05:45	
Air Blank	0.000	05:46	
Diagnostics Check	OK	05:46	

Data Download Date/Time: 07/25/2018 23:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 05:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:07

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 07/08/2018 06:01:06

Diagnostics Check	OK	05:48	
Air Blank	0.000	05:49	
Control Test	0.078	05:49	
Air Blank	0.000	05:49	
Subject Sample	# 1 0.179	05:53	Volume Not Met
Breath Volume	0.156	05:53	
Air Blank	0.000	05:53	
Air Blank	0.000	05:55	
Subject Sample	# 2 0.176	05:56	
Breath Volume	1.929	05:56	
Air Blank	0.000	05:57	
Air Blank	0.000	05:58	
Subject Sample	# 3 0.183	05:59	
Breath Volume	1.578	05:59	
Air Blank	0.000	06:00	
Control Test	0.078	06:00	
Air Blank	0.000	06:01	
Diagnostics Check	OK	06:01	

Data Download Date/Time: 07/25/2018 23:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, WILLIAM, T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: HOOVER'

DATE/TIME OF TEST: 07/09/2018 01:07:05

Diagnostics Check	OK	00:55	
Air Blank	0.000	00:56	
Control Test	0.080	00:56	
Air Blank	0.000	00:56	
Subject Sample	# 1	01:00	No Sample Provided
Breath Volume	0.000	01:00	
Air Blank	0.000	01:00	
Air Blank	0.000	01:02	
Subject Sample	# 2	01:05	No Sample Provided
Breath Volume	0.000	01:05	
Air Blank	0.000	01:06	
Control Test	0.079	01:06	
Air Blank	0.000	01:07	
Diagnostics Check	OK	01:07	

Data Download Date/Time: 07/25/2018 23:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, WILLIAM, T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: LAMBERT

DATE/TIME OF TEST: 07/12/2018 04:33:31

Diagnostics Check	OK	04:22	
Air Blank	0.000	04:22	
Control Test	0.079	04:22	
Air Blank	0.000	04:23	
Subject Sample	# 1	04:26	No Sample Provided
Breath Volume	0.000	04:26	
Air Blank	0.000	04:27	
Air Blank	0.000	04:29	
Subject Sample	# 2	04:32	No Sample Provided
Breath Volume	0.000	04:32	
Air Blank	0.000	04:32	
Control Test	0.080	04:32	
Air Blank	0.000	04:33	
Diagnostics Check	OK	04:33	

Data Download Date/Time: 07/25/2018 23:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: HOPKINS

DATE/TIME OF TEST: 07/14/2018 02:16:01

Diagnostics Check	OK	02:09
Air Blank	0.000	02:10
Control Test	0.079	02:10
Air Blank	0.000	02:10
Subject Sample # 1	0.137	02:11
Breath Volume	1.750	02:11
Air Blank	0.000	02:12
Air Blank	0.000	02:13
Subject Sample # 2	0.135	02:14
Breath Volume	1.437	02:14
Air Blank	0.000	02:15
Control Test	0.080	02:15
Air Blank	0.000	02:15
Diagnostics Check	OK	02:16

Data Download Date/Time: 07/25/2018 23:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REFUSE, RONY,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: REFUSE, RONY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: WALLER

DATE/TIME OF TEST: 07/14/2018 03:38:18

Diagnostics Check	OK	03:30
Air Blank	0.000	03:30
Control Test	0.079	03:31
Air Blank	0.000	03:31
Subject Sample # 1	0.239	03:33
Breath Volume	1.617	03:33
Air Blank	0.000	03:34
Air Blank	0.000	03:36
Subject Sample # 2	0.246	03:36
Breath Volume	2.050	03:36
Air Blank	0.000	03:37
Control Test	0.076	03:37
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

Data Download Date/Time: 07/25/2018 23:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MUNOZ,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: HAAS

DATE/TIME OF TEST: 07/16/2018 00:50:21

Diagnostics Check	OK	00:43
Air Blank	0.000	00:44
Control Test	0.078	00:44
Air Blank	0.000	00:45
Subject Sample # 1	0.110	00:45
Breath Volume	1.121	00:45
Air Blank	0.000	00:46
Air Blank	0.000	00:48
Subject Sample # 2	0.126	00:48
Breath Volume	1.875	00:48
Air Blank	0.000	00:49
Control Test	0.079	00:49
Air Blank	0.000	00:50
Diagnostics Check	OK	00:50

Data Download Date/Time: 07/25/2018 23:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERGER,CHARLES,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BERGER,CHARLES
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: UNGERER

DATE/TIME OF TEST: 07/18/2018 00:58:36

Diagnostics Check	OK	00:50
Air Blank	0.000	00:51
Control Test	0.078	00:51
Air Blank	0.000	00:52
Subject Sample # 1	0.280	00:53
Breath Volume	2.000	00:53
Air Blank	0.000	00:54
Air Blank	0.000	00:56
Subject Sample # 2	0.273	00:57
Breath Volume	2.378	00:57
Air Blank	0.000	00:57
Control Test	0.076	00:58
Air Blank	0.000	00:58
Diagnostics Check	OK	00:58

Data Download Date/Time: 07/25/2018 23:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOODNIGHT,MICHAEL,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GOODNIGHT,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: TARBOX

DATE/TIME OF TEST: 07/18/2018 23:18:32

Diagnostics Check	OK	23:12
Air Blank	0.000	23:12
Control Test	0.078	23:12
Air Blank	0.000	23:13
Subject Sample # 1	0.196	23:13
Breath Volume	2.179	23:13
Air Blank	0.000	23:14
Air Blank	0.000	23:16
Subject Sample # 2	0.201	23:16
Breath Volume	2.597	23:16
Air Blank	0.000	23:17
Control Test	0.078	23:17
Air Blank	0.000	23:18
Diagnostics Check	OK	23:18

Data Download Date/Time: 07/25/2018 23:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: MATKOWSKI

DATE/TIME OF TEST: 07/19/2018 04:45:25

Diagnostics Check	OK	04:38
Air Blank	0.000	04:39
Control Test	0.077	04:39
Air Blank	0.000	04:40
Subject Sample # 1	0.073	04:40
Breath Volume	2.566	04:40
Air Blank	0.000	04:41
Air Blank	0.000	04:43
Subject Sample # 2	0.074	04:43
Breath Volume	3.011	04:43
Air Blank	0.000	04:44
Control Test	0.078	04:44
Air Blank	0.000	04:45
Diagnostics Check	OK	04:45

Data Download Date/Time: 07/25/2018 23:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: CONDELL

DATE/TIME OF TEST: 07/20/2018 02:02:48

Diagnostics Check	OK	01:56
Air Blank	0.000	01:57
Control Test	0.077	01:57
Air Blank	0.000	01:58
Subject Sample # 1	0.000	01:58
Breath Volume	2.117	01:58
Air Blank	0.000	01:59
Air Blank	0.000	02:01
Subject Sample # 2	0.000	02:01
Breath Volume	2.765	02:01
Air Blank	0.000	02:01
Control Test	0.078	02:02
Air Blank	0.000	02:02
Diagnostics Check	OK	02:02

Data Download Date/Time: 07/25/2018 23:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: MERRITT

DATE/TIME OF TEST: 07/20/2018 04:42:05

Diagnostics Check	OK	04:35
Air Blank	0.000	04:36
Control Test	0.078	04:36
Air Blank	0.000	04:36
Subject Sample # 1	0.115	04:37
Breath Volume	1.949	04:37
Air Blank	0.000	04:38
Air Blank	0.000	04:40
Subject Sample # 2	0.108	04:40
Breath Volume	1.332	04:40
Air Blank	0.000	04:41
Control Test	0.079	04:41
Air Blank	0.000	04:42
Diagnostics Check	OK	04:42

Data Download Date/Time: 07/25/2018 23:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI,JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 05:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:55

SUBJECT'S LAST NAME: LADSON

DATE/TIME OF TEST: 07/20/2018 06:21:53

Diagnostics Check	OK	06:15
Air Blank	0.000	06:15
Control Test	0.078	06:16
Air Blank	0.000	06:16
Subject Sample # 1	0.162	06:17
Breath Volume	2.066	06:17
Air Blank	0.000	06:18
Air Blank	0.000	06:19
Subject Sample # 2	0.179	06:20
Breath Volume	2.347	06:20
Air Blank	0.000	06:20
Control Test	0.079	06:21
Air Blank	0.000	06:21
Diagnostics Check	OK	06:21

Data Download Date/Time: 07/25/2018 23:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAIR, RONALD, J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAIR, RONALD
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: MCGUIRE

DATE/TIME OF TEST: 07/21/2018 00:37:10

Diagnostics Check	OK	00:31
Air Blank	0.000	00:31
Control Test	0.079	00:31
Air Blank	0.000	00:32
Subject Sample # 1	0.196	00:32
Breath Volume	2.476	00:32
Air Blank	0.000	00:33
Air Blank	0.000	00:35
Subject Sample # 2	0.197	00:35
Breath Volume	2.472	00:35
Air Blank	0.000	00:36
Control Test	0.079	00:36
Air Blank	0.000	00:37
Diagnostics Check	OK	00:37

Data Download Date/Time: 07/25/2018 23:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: MAJERCIK

DATE/TIME OF TEST: 07/21/2018 01:21:58

Diagnostics Check	OK	01:18	
Air Blank	0.000	01:19	
Control Test	0.076	01:19	
Air Blank	0.000	01:19	
Subject Sample	# 1	01:20	Subject Test Refused
Breath Volume	0.000	01:20	
Air Blank	0.000	01:21	
Control Test	0.077	01:21	
Air Blank	0.000	01:21	
Diagnostics Check	OK	01:21	

Data Download Date/Time: 07/25/2018 23:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERGER,CHARLES,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BERGER,CHARLES
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: HRABOVSKY

DATE/TIME OF TEST: 07/21/2018 02:20:47

Diagnostics Check	OK	02:16	
Air Blank	0.000	02:17	
Control Test	0.079	02:17	
Air Blank	0.000	02:18	
Subject Sample	# 1	02:19	Subject Test Refused
Breath Volume	0.000	02:19	
Air Blank	0.000	02:19	
Control Test	0.077	02:20	
Air Blank	0.000	02:20	
Diagnostics Check	OK	02:20	

Data Download Date/Time: 07/25/2018 23:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HART,RANDALL,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HART,RANDALL
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: BUI
ARREST TIME: 20:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: PAUL

DATE/TIME OF TEST: 07/22/2018 22:06:23

Diagnostics Check	OK	21:55	
Air Blank	0.000	21:55	
Control Test	0.078	21:55	
Air Blank	0.000	21:56	
Subject Sample	# 1	21:59	No Sample Provided
Breath Volume	0.000	21:59	
Air Blank	0.000	21:59	
Air Blank	0.000	22:01	
Subject Sample	# 2	22:05	No Sample Provided
Breath Volume	0.000	22:05	
Air Blank	0.000	22:05	
Control Test	0.078	22:05	
Air Blank	0.000	22:06	
Diagnostics Check	OK	22:06	

Data Download Date/Time: 07/25/2018 23:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE,DEPUTY WILLIAM,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WILTSE,DEPUTY WILLIAM
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: DISNEY

DATE/TIME OF TEST: 07/25/2018 00:50:52

Diagnostics Check	OK	00:44
Air Blank	0.000	00:44
Control Test	0.078	00:45
Air Blank	0.000	00:45
Subject Sample # 1	0.009	00:46
Breath Volume	2.738	00:46
Air Blank	0.000	00:46
Air Blank	0.000	00:48
Subject Sample # 2	0.009	00:49
Breath Volume	2.480	00:49
Air Blank	0.000	00:49
Control Test	0.079	00:50
Air Blank	0.000	00:50
Diagnostics Check	OK	00:50

Data Download Date/Time: 07/25/2018 23:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE,DEPUTY WILLIAM,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RODGERS,DEPUTY R.
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 05:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:34

SUBJECT'S LAST NAME: PLANTE II

DATE/TIME OF TEST: 07/25/2018 06:02:39

Diagnostics Check	OK	05:55
Air Blank	0.000	05:55
Control Test	0.078	05:55
Air Blank	0.000	05:56
Subject Sample # 1	0.078	05:58
Breath Volume	1.894	05:58
Air Blank	0.000	05:58
Air Blank	0.000	06:00
Subject Sample # 2	0.069	06:01
Breath Volume	1.683	06:01
Air Blank	0.000	06:01
Control Test	0.079	06:02
Air Blank	0.000	06:02
Diagnostics Check	OK	06:02

Data Download Date/Time: 07/25/2018 23:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REFUSE, RONY,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: REFUSE, RONY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: SHARKEY

DATE/TIME OF TEST: 07/26/2018 04:44:42

Diagnostics Check	OK	04:41	
Air Blank	0.000	04:42	
Control Test	0.078	04:42	
Air Blank	0.000	04:43	
Subject Sample	# 1	04:43	Subject Test Refused
Breath Volume	0.000	04:43	
Air Blank	0.000	04:43	
Control Test	0.078	04:44	
Air Blank	0.000	04:44	
Diagnostics Check	OK	04:44	

Data Download Date/Time: 08/02/2018 22:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REFUSE, RONY,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: REFUSE, RONY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:25

SUBJECT'S LAST NAME: SHARKEY

DATE/TIME OF TEST: 07/26/2018 05:49:44

Diagnostics Check	OK	05:46	
Air Blank	0.000	05:46	
Control Test	0.078	05:47	
Air Blank	0.000	05:47	
Subject Sample	# 1	05:48	Subject Test Refused
Breath Volume	0.000	05:48	
Air Blank	0.000	05:48	
Control Test	0.079	05:49	
Air Blank	0.000	05:49	
Diagnostics Check	OK	05:49	

Data Download Date/Time: 08/02/2018 22:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONGMIRE,BENJAMIN,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: LONGMIRE,BENJAMIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: MATHER

DATE/TIME OF TEST: 07/26/2018 20:49:49

Diagnostics Check	OK	20:45	
Air Blank	0.000	20:46	
Control Test	0.076	20:46	
Air Blank	0.000	20:47	
Subject Sample	# 1	20:48	Subject Test Refused
Breath Volume	0.000	20:48	
Air Blank	0.000	20:48	
Control Test	0.078	20:49	
Air Blank	0.000	20:49	
Diagnostics Check	OK	20:49	

Data Download Date/Time: 08/02/2018 22:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REFUSE, RONY,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: REFUSE, RONY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: WEMMER

DATE/TIME OF TEST: 07/29/2018 03:56:00

Diagnostics Check	OK	03:53	
Air Blank	0.000	03:53	
Control Test	0.078	03:53	
Air Blank	0.000	03:54	
Subject Sample	# 1	03:54	Subject Test Refused
Breath Volume	0.000	03:54	
Air Blank	0.000	03:55	
Control Test	0.077	03:55	
Air Blank	0.000	03:55	
Diagnostics Check	OK	03:56	

Data Download Date/Time: 08/02/2018 22:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROSS,MARK,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GEOGHEGAN,LEAH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: CARDENAS ALVAREZ

DATE/TIME OF TEST: 07/30/2018 03:23:28

Diagnostics Check	OK	03:20	
Air Blank	0.000	03:20	
Control Test	0.078	03:20	
Air Blank	0.000	03:21	
Subject Sample # 1	0.000	03:21	
Breath Volume	2.894	03:21	
Air Blank		03:22	RFI Detect
Air Blank	0.000	03:22	
Control Test	0.078	03:22	
Air Blank	0.000	03:23	
Diagnostics Check	OK	03:23	

Data Download Date/Time: 08/02/2018 22:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROSS,MARK,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GEOHEGAN,LEAH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: CARDENAS ALVAREZ

DATE/TIME OF TEST: 07/30/2018 03:30:22

Diagnostics Check	OK	03:27	
Air Blank	0.000	03:27	
Control Test	0.078	03:28	
Air Blank	0.000	03:28	
Subject Sample	# 1	03:29	RFI Detect
Breath Volume	0.000	03:29	
Air Blank	0.000	03:29	
Control Test	0.078	03:29	
Air Blank	0.000	03:30	
Diagnostics Check	OK	03:30	

Data Download Date/Time: 08/02/2018 22:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLE,S,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HOLE,S
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: GUY

DATE/TIME OF TEST: 07/01/2018 01:04:19

Diagnostics Check	OK	00:57
Air Blank	0.000	00:58
Control Test	0.077	00:58
Air Blank	0.000	00:59
Subject Sample # 1	0.166	00:59
Breath Volume	1.250	00:59
Air Blank	0.000	01:00
Air Blank	0.000	01:02
Subject Sample # 2	0.179	01:02
Breath Volume	1.214	01:02
Air Blank	0.000	01:03
Control Test	0.077	01:03
Air Blank	0.000	01:04
Diagnostics Check	OK	01:04

Data Download Date/Time: 07/25/2018 22:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMIH, WILLIAM, T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: MCCUSKER

DATE/TIME OF TEST: 07/07/2018 01:28:17

Diagnostics Check	OK	01:15	
Air Blank	0.000	01:15	
Control Test	0.078	01:15	
Air Blank	0.000	01:16	
Subject Sample	# 1 0.000	01:19	Volume Not Met
Breath Volume	0.472	01:19	
Air Blank	0.000	01:19	
Air Blank	0.000	01:21	
Subject Sample	# 2 0.000	01:22	
Breath Volume	2.062	01:22	
Air Blank	0.000	01:23	
Air Blank	0.000	01:25	
Subject Sample	# 3 0.000	01:26	
Breath Volume	1.566	01:26	
Air Blank	0.000	01:27	
Control Test	0.079	01:27	
Air Blank	0.000	01:28	
Diagnostics Check	OK	01:28	

Data Download Date/Time: 07/25/2018 22:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, WILLIAM, T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: ROSADO

DATE/TIME OF TEST: 07/07/2018 04:38:54

Diagnostics Check	OK	04:26	
Air Blank	0.000	04:27	
Control Test	0.078	04:27	
Air Blank	0.000	04:28	
Subject Sample	# 1	04:31	No Sample Provided
Breath Volume	0.000	04:31	
Air Blank	0.000	04:31	
Air Blank	0.000	04:33	
Subject Sample	# 2	0.118	04:34
Breath Volume	2.804	04:34	
Air Blank	0.000	04:34	
Air Blank	0.000	04:36	
Subject Sample	# 3	0.123	04:37
Breath Volume	3.082	04:37	
Air Blank	0.000	04:38	
Control Test	0.078	04:38	
Air Blank	0.000	04:38	
Diagnostics Check	OK	04:38	

Data Download Date/Time: 07/25/2018 22:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHRISTIE,MATT,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHRISTIE,MATT
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: FERREIRA

DATE/TIME OF TEST: 07/12/2018 03:51:14

Diagnostics Check	OK	03:48	
Air Blank	0.000	03:48	
Control Test	0.078	03:49	
Air Blank	0.000	03:49	
Subject Sample	# 1	03:49	Subject Test Refused
Breath Volume	0.000	03:49	
Air Blank	0.000	03:50	
Control Test	0.077	03:50	
Air Blank	0.000	03:51	
Diagnostics Check	OK	03:51	

Data Download Date/Time: 07/25/2018 22:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

INSTRUMENT REGISTERED TO: INDIALANTIC P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,TIMOTHY,C
BREATH TEST OPERATOR AGENCY: Indialantic PD
ARRESTING OFFICER: VILA,ALEXANDER
ARRESTING OFFICER AGENCY: Melbourne Beach PD
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 09:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:20

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 07/03/2018 09:49:30

Diagnostics Check	OK	09:43
Air Blank	0.000	09:43
Control Test	0.079	09:44
Air Blank	0.000	09:44
Subject Sample # 1	0.000	09:45
Breath Volume	2.980	09:45
Air Blank	0.000	09:45
Air Blank	0.000	09:47
Subject Sample # 2	0.000	09:48
Breath Volume	3.339	09:48
Air Blank	0.000	09:48
Control Test	0.079	09:48
Air Blank	0.000	09:49
Diagnostics Check	OK	09:49

Data Download Date/Time: 07/28/2018 07:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKSON,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: JACKSON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 06/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 09:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:15

SUBJECT'S LAST NAME: DOMINGUEZ

DATE/TIME OF TEST: 07/04/2018 10:06:01

Diagnostics Check	OK	10:00
Air Blank	0.000	10:00
Control Test	0.082	10:00
Air Blank	0.000	10:01
Subject Sample # 1	0.000	10:01
Breath Volume	3.453	10:01
Air Blank	0.000	10:02
Air Blank	0.000	10:04
Subject Sample # 2	0.000	10:04
Breath Volume	3.664	10:04
Air Blank	0.000	10:05
Control Test	0.079	10:05
Air Blank	0.000	10:05
Diagnostics Check	OK	10:06

Data Download Date/Time: 07/08/2018 19:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER, ADAM, H
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: POTTER, ADAM
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: JEREZ

DATE/TIME OF TEST: 07/06/2018 03:28:01

Diagnostics Check	OK	03:21
Air Blank	0.000	03:22
Control Test	0.082	03:22
Air Blank	0.000	03:23
Subject Sample # 1	0.109	03:23
Breath Volume	2.097	03:23
Air Blank	0.000	03:24
Air Blank	0.000	03:26
Subject Sample # 2	0.100	03:26
Breath Volume	2.425	03:26
Air Blank	0.000	03:27
Control Test	0.082	03:27
Air Blank	0.000	03:27
Diagnostics Check	OK	03:28

Data Download Date/Time: 07/08/2018 19:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKSON,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: JACKSON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 07/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 19:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: LAUX

DATE/TIME OF TEST: 07/10/2018 20:34:50

Diagnostics Check	OK	20:28
Air Blank	0.000	20:28
Control Test	0.081	20:29
Air Blank	0.000	20:29
Subject Sample # 1	0.203	20:30
Breath Volume	3.113	20:30
Air Blank	0.000	20:30
Air Blank	0.000	20:32
Subject Sample # 2	0.201	20:33
Breath Volume	3.035	20:33
Air Blank	0.000	20:33
Control Test	0.083	20:34
Air Blank	0.000	20:34
Diagnostics Check	OK	20:34

Data Download Date/Time: 08/22/2018 16:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAGER,KEVIN,M
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: NA,NA
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 07/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: OTHER
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: WEBER

DATE/TIME OF TEST: 07/18/2018 04:47:43

Diagnostics Check	OK	04:41
Air Blank	0.000	04:42
Control Test	0.082	04:42
Air Blank	0.000	04:43
Subject Sample # 1	0.000	04:43
Breath Volume	2.238	04:43
Air Blank	0.000	04:44
Air Blank	0.000	04:46
Subject Sample # 2	0.000	04:46
Breath Volume	2.472	04:46
Air Blank	0.000	04:46
Control Test	0.082	04:47
Air Blank	0.000	04:47
Diagnostics Check	OK	04:47

Data Download Date/Time: 08/22/2018 16:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCKER, WILLIAM,
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: TUCKER, WILLIAM
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 07/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: LUCAS MARTINEZ

DATE/TIME OF TEST: 07/28/2018 23:55:32

Diagnostics Check	OK	23:52	
Air Blank	0.000	23:53	
Control Test	0.082	23:53	
Air Blank	0.000	23:53	
Subject Sample	# 1	23:54	Subject Test Refused
Breath Volume	0.000	23:54	
Air Blank	0.000	23:54	
Control Test	0.083	23:54	
Air Blank	0.000	23:55	
Diagnostics Check	OK	23:55	

Data Download Date/Time: 08/22/2018 16:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MORGAN,D,

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: REARDON,J

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 26316080A2

DRY GAS EXPIRY DATE: 11/05/2018

VIOLATION CODE: DUI

ARREST TIME: 07:16

GENDER: Male

OBSERVATION PERIOD BEGAN: 07:21

SUBJECT'S LAST NAME: CAMPBELL ARTHUR

DATE/TIME OF TEST: 07/01/2018 07:56:27

Diagnostics Check	OK	07:50
Air Blank	0.000	07:50
Control Test	0.079	07:50
Air Blank	0.000	07:51
Subject Sample # 1	0.122	07:51
Breath Volume	2.144	07:51
Air Blank	0.000	07:52
Air Blank	0.000	07:54
Subject Sample # 2	0.119	07:54
Breath Volume	2.699	07:54
Air Blank	0.000	07:55
Control Test	0.078	07:55
Air Blank	0.000	07:56
Diagnostics Check	OK	07:56

Data Download Date/Time: 07/26/2018 19:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FREY,ROBERT,P

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LOGAN,J

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 26316080A2

DRY GAS EXPIRY DATE: 11/05/2018

VIOLATION CODE: DUI

ARREST TIME: 03:14

GENDER: Female

OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: OSWALT

DATE/TIME OF TEST: 07/04/2018 04:42:10

Diagnostics Check	OK	04:30	
Air Blank	0.000	04:30	
Control Test	0.080	04:31	
Air Blank	0.000	04:31	
Subject Sample	# 1 0.176	04:34	Volume Not Met
Breath Volume	0.906	04:34	
Air Blank	0.000	04:35	
Air Blank	0.000	04:37	
Subject Sample	# 2 0.187	04:37	
Breath Volume	1.507	04:37	
Air Blank	0.000	04:38	
Air Blank	0.000	04:40	
Subject Sample	# 3 0.182	04:40	
Breath Volume	1.574	04:40	
Air Blank	0.000	04:41	
Control Test	0.079	04:41	
Air Blank	0.000	04:42	
Diagnostics Check	OK	04:42	

Data Download Date/Time: 07/26/2018 19:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VURRO,MARC,E
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: MILIAN,SHAWN
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 07/04/2018 04:58:14

Diagnostics Check	OK	04:51
Air Blank	0.000	04:52
Control Test	0.080	04:52
Air Blank	0.000	04:52
Subject Sample # 1	0.117	04:53
Breath Volume	2.187	04:53
Air Blank	0.000	04:54
Air Blank	0.000	04:56
Subject Sample # 2	0.120	04:56
Breath Volume	1.500	04:56
Air Blank	0.000	04:57
Control Test	0.079	04:57
Air Blank	0.000	04:58
Diagnostics Check	OK	04:58

Data Download Date/Time: 07/26/2018 19:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HAMILTON,J,W

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: HAMILTON,J

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 26316080A2

DRY GAS EXPIRY DATE: 11/05/2018

VIOLATION CODE: DUI

ARREST TIME: 23:15

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: DUNCAN

DATE/TIME OF TEST: 07/17/2018 00:19:11

Diagnostics Check	OK	00:10
Air Blank	0.000	00:10
Control Test	0.079	00:11
Air Blank	0.000	00:11
Subject Sample # 1	0.181	00:13
Breath Volume	1.679	00:13
Air Blank	0.000	00:14
Air Blank	0.000	00:16
Subject Sample # 2	0.190	00:17
Breath Volume	1.996	00:17
Air Blank	0.000	00:18
Control Test	0.079	00:18
Air Blank	0.000	00:19
Diagnostics Check	OK	00:19

Data Download Date/Time: 07/26/2018 19:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MORGAN,D,

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: REARDON,J

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 26316080A2

DRY GAS EXPIRY DATE: 11/05/2018

VIOLATION CODE: DUI

ARREST TIME: 07:31

GENDER: Female

OBSERVATION PERIOD BEGAN: 10:50

SUBJECT'S LAST NAME: REDDING

DATE/TIME OF TEST: 07/18/2018 11:24:26

Diagnostics Check	OK	11:18
Air Blank	0.000	11:18
Control Test	0.080	11:18
Air Blank	0.000	11:19
Subject Sample # 1	0.137	11:20
Breath Volume	1.437	11:20
Air Blank	0.000	11:20
Air Blank	0.000	11:22
Subject Sample # 2	0.133	11:22
Breath Volume	1.652	11:22
Air Blank	0.000	11:23
Control Test	0.079	11:23
Air Blank	0.000	11:24
Diagnostics Check	OK	11:24

Data Download Date/Time: 07/26/2018 19:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ROLPH,AMANDA,L

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: ROLPH,AMANDA

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 26316080A2

DRY GAS EXPIRY DATE: 11/05/2018

VIOLATION CODE: DUI

ARREST TIME: 01:33

GENDER: Female

OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: BENJAMIN

DATE/TIME OF TEST: 07/20/2018 02:25:09

Diagnostics Check	OK	02:17
Air Blank	0.000	02:18
Control Test	0.079	02:18
Air Blank	0.000	02:18
Subject Sample # 1	0.167	02:20
Breath Volume	2.203	02:20
Air Blank	0.000	02:21
Air Blank	0.000	02:22
Subject Sample # 2	0.168	02:23
Breath Volume	1.914	02:23
Air Blank	0.000	02:24
Control Test	0.078	02:24
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

Data Download Date/Time: 07/26/2018 19:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ROLPH,AMANDA,L

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: WHITE,TAYLOR

ARRESTING OFFICER AGENCY: Melbourne PD

LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 26316080A2

DRY GAS EXPIRY DATE: 11/05/2018

VIOLATION CODE: DUI

ARREST TIME: 00:36

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: JAMES

DATE/TIME OF TEST: 07/20/2018 03:12:35

Diagnostics Check	OK	03:06
Air Blank	0.000	03:06
Control Test	0.079	03:07
Air Blank	0.000	03:07
Subject Sample # 1	0.093	03:08
Breath Volume	3.105	03:08
Air Blank	0.000	03:08
Air Blank	0.000	03:10
Subject Sample # 2	0.089	03:11
Breath Volume	3.039	03:11
Air Blank	0.000	03:11
Control Test	0.078	03:11
Air Blank	0.000	03:12
Diagnostics Check	OK	03:12

Data Download Date/Time: 07/26/2018 19:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FREY,ROBERT,P

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: BUCKLIN,SHANE

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 26316080A2

DRY GAS EXPIRY DATE: 11/05/2018

VIOLATION CODE: DUI

ARREST TIME: 04:00

GENDER: Female

OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: MEJIA-BLEN

DATE/TIME OF TEST: 07/21/2018 05:04:31

Diagnostics Check	OK	04:56
Air Blank	0.000	04:56
Control Test	0.080	04:57
Air Blank	0.000	04:57
Subject Sample # 1	0.128	04:59
Breath Volume	1.488	04:59
Air Blank	0.000	05:00
Air Blank	0.000	05:02
Subject Sample # 2	0.131	05:02
Breath Volume	1.562	05:02
Air Blank	0.000	05:03
Control Test	0.078	05:03
Air Blank	0.000	05:04
Diagnostics Check	OK	05:04

Data Download Date/Time: 07/26/2018 19:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HAMILTON,J,W

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: HAMILTON,J

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 26316080A2

DRY GAS EXPIRY DATE: 11/05/2018

VIOLATION CODE: DUI

ARREST TIME: 22:58

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: MALINOWSKI

DATE/TIME OF TEST: 07/25/2018 23:51:10

Diagnostics Check	OK	23:44
Air Blank	0.000	23:45
Control Test	0.080	23:45
Air Blank	0.000	23:45
Subject Sample # 1	0.176	23:46
Breath Volume	3.820	23:46
Air Blank	0.000	23:47
Air Blank	0.000	23:49
Subject Sample # 2	0.170	23:49
Breath Volume	3.062	23:49
Air Blank	0.000	23:50
Control Test	0.079	23:50
Air Blank	0.000	23:51
Diagnostics Check	OK	23:51

Data Download Date/Time: 07/26/2018 19:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD, WILLIAM,
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: HOWARD, WILLIAM
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 07/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 07/28/2018 20:14:32

Diagnostics Check	OK	20:08
Air Blank	0.000	20:08
Control Test	0.080	20:08
Air Blank	0.000	20:09
Subject Sample # 1	0.245	20:09
Breath Volume	1.773	20:09
Air Blank	0.000	20:10
Air Blank	0.000	20:12
Subject Sample # 2	0.238	20:12
Breath Volume	1.644	20:12
Air Blank	0.000	20:13
Control Test	0.080	20:13
Air Blank	0.000	20:14
Diagnostics Check	OK	20:14

Data Download Date/Time: 08/28/2018 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: GRIFFITH

DATE/TIME OF TEST: 07/06/2018 01:54:51

Diagnostics Check	OK	01:43	
Air Blank	0.000	01:43	
Control Test	0.081	01:44	
Air Blank	0.000	01:44	
Subject Sample	# 1	01:47	No Sample Provided
Breath Volume	0.000	01:47	
Air Blank	0.000	01:48	
Air Blank	0.000	01:50	
Subject Sample	# 2	01:53	No Sample Provided
Breath Volume	0.000	01:53	
Air Blank	0.000	01:53	
Control Test	0.081	01:54	
Air Blank	0.000	01:54	
Diagnostics Check	OK	01:54	

Data Download Date/Time: 07/06/2018 12:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SORRELLS,M
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 07/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: HAHN

DATE/TIME OF TEST: 07/06/2018 04:59:53

Diagnostics Check	OK	04:52
Air Blank	0.000	04:52
Control Test	0.081	04:53
Air Blank	0.000	04:53
Subject Sample # 1	0.108	04:54
Breath Volume	1.195	04:54
Air Blank	0.000	04:55
Air Blank	0.000	04:57
Subject Sample # 2	0.119	04:58
Breath Volume	1.218	04:58
Air Blank	0.000	04:58
Control Test	0.079	04:59
Air Blank	0.000	04:59
Diagnostics Check	OK	04:59

Data Download Date/Time: 07/06/2018 12:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY, SHANON, E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRYANT, R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: RAY

DATE/TIME OF TEST: 07/08/2018 02:35:51

Diagnostics Check	OK	02:32	
Air Blank	0.000	02:33	
Control Test	0.080	02:33	
Air Blank	0.000	02:34	
Subject Sample	# 1	02:34	Subject Test Refused
Breath Volume	0.000	02:34	
Air Blank	0.000	02:34	
Control Test	0.080	02:35	
Air Blank	0.000	02:35	
Diagnostics Check	OK	02:35	

Data Download Date/Time: 07/18/2018 14:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY, SHANON, E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: COWAN, B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: SAVIORY

DATE/TIME OF TEST: 07/08/2018 04:35:52

Diagnostics Check	OK	04:32	
Air Blank	0.000	04:33	
Control Test	0.081	04:33	
Air Blank	0.000	04:34	
Subject Sample	# 1	04:34	Subject Test Refused
Breath Volume	0.000	04:34	
Air Blank	0.000	04:34	
Control Test	0.080	04:35	
Air Blank	0.000	04:35	
Diagnostics Check	OK	04:35	

Data Download Date/Time: 07/18/2018 14:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MATHIS,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: MULERO

DATE/TIME OF TEST: 07/10/2018 03:02:26

Diagnostics Check	OK	02:59	
Air Blank	0.000	02:59	
Control Test	0.080	03:00	
Air Blank	0.000	03:00	
Subject Sample	# 1	03:01	Subject Test Refused
Breath Volume	0.000	03:01	
Air Blank	0.000	03:01	
Control Test	0.081	03:01	
Air Blank	0.000	03:02	
Diagnostics Check	OK	03:02	

Data Download Date/Time: 07/18/2018 14:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: BAJALIA

DATE/TIME OF TEST: 07/11/2018 00:53:47

Diagnostics Check	OK	00:47
Air Blank	0.000	00:47
Control Test	0.081	00:48
Air Blank	0.000	00:48
Subject Sample # 1	0.000	00:49
Breath Volume	3.015	00:49
Air Blank	0.000	00:49
Air Blank	0.000	00:51
Subject Sample # 2	0.000	00:52
Breath Volume	3.910	00:52
Air Blank	0.000	00:52
Control Test	0.080	00:53
Air Blank	0.000	00:53
Diagnostics Check	OK	00:53

Data Download Date/Time: 07/18/2018 14:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MARRERO,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: FAGAN

DATE/TIME OF TEST: 07/11/2018 02:01:40

Diagnostics Check	OK	01:58	
Air Blank	0.000	01:59	
Control Test	0.081	01:59	
Air Blank	0.000	02:00	
Subject Sample	# 1	02:00	Subject Test Refused
Breath Volume	0.000	02:00	
Air Blank	0.000	02:00	
Control Test	0.080	02:01	
Air Blank	0.000	02:01	
Diagnostics Check	OK	02:01	

Data Download Date/Time: 07/18/2018 14:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS, CHRISTOPHER, D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRYANT, R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 07/11/2018 04:05:33

Diagnostics Check	OK	03:59
Air Blank	0.000	03:59
Control Test	0.081	04:00
Air Blank	0.000	04:00
Subject Sample # 1	0.174	04:00
Breath Volume	1.656	04:00
Air Blank	0.000	04:01
Air Blank	0.000	04:03
Subject Sample # 2	0.174	04:03
Breath Volume	1.910	04:03
Air Blank	0.000	04:04
Control Test	0.080	04:04
Air Blank	0.000	04:05
Diagnostics Check	OK	04:05

Data Download Date/Time: 07/18/2018 14:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MARRERO,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: BOND

DATE/TIME OF TEST: 07/12/2018 01:45:05

Diagnostics Check	OK	01:42	
Air Blank	0.000	01:42	
Control Test	0.080	01:42	
Air Blank	0.000	01:43	
Subject Sample	# 1	01:43	Subject Test Refused
Breath Volume	0.000	01:43	
Air Blank	0.000	01:44	
Control Test	0.080	01:44	
Air Blank	0.000	01:45	
Diagnostics Check	OK	01:45	

Data Download Date/Time: 07/18/2018 14:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: ABUALGANAM

DATE/TIME OF TEST: 07/12/2018 03:55:41

Diagnostics Check	OK	03:48
Air Blank	0.000	03:49
Control Test	0.081	03:49
Air Blank	0.000	03:50
Subject Sample # 1	0.088	03:50
Breath Volume	2.007	03:50
Air Blank	0.000	03:51
Air Blank	0.000	03:53
Subject Sample # 2	0.081	03:54
Breath Volume	1.664	03:54
Air Blank	0.000	03:54
Control Test	0.080	03:55
Air Blank	0.000	03:55
Diagnostics Check	OK	03:55

Data Download Date/Time: 07/18/2018 14:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: SIMON

DATE/TIME OF TEST: 07/13/2018 02:11:54

Diagnostics Check	OK	02:05
Air Blank	0.000	02:06
Control Test	0.080	02:06
Air Blank	0.000	02:07
Subject Sample # 1	0.011	02:07
Breath Volume	3.445	02:07
Air Blank	0.000	02:08
Air Blank	0.000	02:10
Subject Sample # 2	0.011	02:10
Breath Volume	3.035	02:10
Air Blank	0.000	02:10
Control Test	0.081	02:11
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

Data Download Date/Time: 07/18/2018 14:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: NEAL

DATE/TIME OF TEST: 07/13/2018 04:11:21

Diagnostics Check	OK	04:04
Air Blank	0.000	04:05
Control Test	0.081	04:05
Air Blank	0.000	04:05
Subject Sample # 1	0.198	04:06
Breath Volume	2.527	04:06
Air Blank	0.000	04:07
Air Blank	0.000	04:09
Subject Sample # 2	0.195	04:09
Breath Volume	2.292	04:09
Air Blank	0.000	04:10
Control Test	0.080	04:10
Air Blank	0.000	04:11
Diagnostics Check	OK	04:11

Data Download Date/Time: 07/18/2018 14:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: RIGBY

DATE/TIME OF TEST: 07/13/2018 23:45:08

Diagnostics Check	OK	23:38
Air Blank	0.000	23:38
Control Test	0.080	23:38
Air Blank	0.000	23:39
Subject Sample # 1	0.075	23:40
Breath Volume	2.769	23:40
Air Blank	0.000	23:41
Air Blank	0.000	23:42
Subject Sample # 2	0.078	23:43
Breath Volume	2.789	23:43
Air Blank	0.000	23:44
Control Test	0.080	23:44
Air Blank	0.000	23:45
Diagnostics Check	OK	23:45

Data Download Date/Time: 07/18/2018 14:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 07/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: MITCHNER

DATE/TIME OF TEST: 07/14/2018 04:02:20

Diagnostics Check	OK	03:56
Air Blank	0.000	03:56
Control Test	0.080	03:56
Air Blank	0.000	03:57
Subject Sample # 1	0.076	03:57
Breath Volume	3.339	03:57
Air Blank	0.000	03:58
Air Blank	0.000	04:00
Subject Sample # 2	0.080	04:00
Breath Volume	3.531	04:00
Air Blank	0.000	04:01
Control Test	0.080	04:01
Air Blank	0.000	04:02
Diagnostics Check	OK	04:02

Data Download Date/Time: 07/18/2018 14:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: YANES DOMINGUEZ

DATE/TIME OF TEST: 07/15/2018 00:42:40

Diagnostics Check	OK	00:35
Air Blank	0.000	00:36
Control Test	0.080	00:36
Air Blank	0.000	00:37
Subject Sample # 1	0.182	00:38
Breath Volume	1.890	00:38
Air Blank	0.000	00:38
Air Blank	0.000	00:40
Subject Sample # 2	0.181	00:40
Breath Volume	2.980	00:40
Air Blank	0.000	00:41
Control Test	0.079	00:42
Air Blank	0.000	00:42
Diagnostics Check	OK	00:42

Data Download Date/Time: 07/18/2018 14:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: FOREST

DATE/TIME OF TEST: 07/15/2018 03:41:17

Diagnostics Check	OK	03:35
Air Blank	0.000	03:35
Control Test	0.081	03:35
Air Blank	0.000	03:36
Subject Sample # 1	0.024	03:36
Breath Volume	2.890	03:36
Air Blank	0.000	03:37
Air Blank	0.000	03:39
Subject Sample # 2	0.026	03:39
Breath Volume	2.761	03:39
Air Blank	0.000	03:40
Control Test	0.081	03:40
Air Blank	0.000	03:41
Diagnostics Check	OK	03:41

Data Download Date/Time: 07/18/2018 14:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS, CHRISTOPHER, D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY, T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: VESTAL

DATE/TIME OF TEST: 07/16/2018 01:56:12

Diagnostics Check	OK	01:49
Air Blank	0.000	01:50
Control Test	0.080	01:50
Air Blank	0.000	01:51
Subject Sample # 1	0.163	01:51
Breath Volume	2.769	01:51
Air Blank	0.000	01:52
Air Blank	0.000	01:54
Subject Sample # 2	0.154	01:54
Breath Volume	1.578	01:54
Air Blank	0.000	01:55
Control Test	0.080	01:55
Air Blank	0.000	01:56
Diagnostics Check	OK	01:56

Data Download Date/Time: 07/18/2018 14:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKNINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: D'ANTONI

DATE/TIME OF TEST: 07/16/2018 04:33:12

Diagnostics Check	OK	04:26
Air Blank	0.000	04:27
Control Test	0.081	04:27
Air Blank	0.000	04:27
Subject Sample # 1	0.151	04:28
Breath Volume	3.257	04:28
Air Blank	0.000	04:29
Air Blank	0.000	04:30
Subject Sample # 2	0.152	04:31
Breath Volume	3.339	04:31
Air Blank	0.000	04:32
Control Test	0.079	04:32
Air Blank	0.000	04:33
Diagnostics Check	OK	04:33

Data Download Date/Time: 07/18/2018 14:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HOOVER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 07/17/2018 02:59:32

Diagnostics Check	OK	02:56	
Air Blank	0.000	02:57	
Control Test	0.080	02:57	
Air Blank	0.000	02:57	
Subject Sample	# 1	02:58	Subject Test Refused
Breath Volume	0.000	02:58	
Air Blank	0.000	02:58	
Control Test	0.080	02:58	
Air Blank	0.000	02:59	
Diagnostics Check	OK	02:59	

Data Download Date/Time: 07/18/2018 14:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 07/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: SLOAN

DATE/TIME OF TEST: 07/17/2018 04:19:15

Diagnostics Check	OK	04:16	
Air Blank	0.000	04:16	
Control Test	0.080	04:17	
Air Blank	0.000	04:17	
Subject Sample	# 1	04:17	Subject Test Refused
Breath Volume	0.000	04:17	
Air Blank	0.000	04:18	
Control Test	0.080	04:18	
Air Blank	0.000	04:19	
Diagnostics Check	OK	04:19	

Data Download Date/Time: 07/18/2018 14:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILBERGER,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MCCRORY,D
ARRESTING OFFICER AGENCY: Jax Airport Auth
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 12:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:45

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 07/18/2018 15:09:34

Diagnostics Check	OK	15:06	
Air Blank	0.000	15:06	
Control Test	0.081	15:07	
Air Blank	0.000	15:07	
Subject Sample	# 1	15:08	Subject Test Refused
Breath Volume	0.000	15:08	
Air Blank	0.000	15:08	
Control Test	0.080	15:08	
Air Blank	0.000	15:09	
Diagnostics Check	OK	15:09	

Data Download Date/Time: 08/17/2018 11:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: LANES

DATE/TIME OF TEST: 07/19/2018 02:55:17

Diagnostics Check	OK	02:49
Air Blank	0.000	02:49
Control Test	0.080	02:49
Air Blank	0.000	02:50
Subject Sample # 1	0.131	02:50
Breath Volume	1.726	02:50
Air Blank	0.000	02:51
Air Blank	0.000	02:53
Subject Sample # 2	0.126	02:53
Breath Volume	2.804	02:53
Air Blank	0.000	02:54
Control Test	0.079	02:54
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

Data Download Date/Time: 08/17/2018 11:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL, EDWARD, A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FILBERT, E
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: MANNING

DATE/TIME OF TEST: 07/20/2018 02:55:27

Diagnostics Check	OK	02:48
Air Blank	0.000	02:49
Control Test	0.080	02:49
Air Blank	0.000	02:50
Subject Sample # 1	0.213	02:50
Breath Volume	2.285	02:50
Air Blank	0.000	02:51
Air Blank	0.000	02:53
Subject Sample # 2	0.196	02:53
Breath Volume	2.089	02:53
Air Blank	0.000	02:54
Control Test	0.079	02:54
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

Data Download Date/Time: 08/17/2018 11:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BOWERS,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: THERMITUS

DATE/TIME OF TEST: 07/20/2018 06:07:14

Diagnostics Check	OK	06:01
Air Blank	0.000	06:01
Control Test	0.080	06:01
Air Blank	0.000	06:02
Subject Sample # 1	0.099	06:02
Breath Volume	2.195	06:02
Air Blank	0.000	06:03
Air Blank	0.000	06:05
Subject Sample # 2	0.095	06:05
Breath Volume	1.914	06:05
Air Blank	0.000	06:06
Control Test	0.080	06:06
Air Blank	0.000	06:07
Diagnostics Check	OK	06:07

Data Download Date/Time: 08/17/2018 11:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,J,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 07/21/2018 21:13:11

Diagnostics Check	OK	21:01	
Air Blank	0.000	21:02	
Control Test	0.080	21:02	
Air Blank	0.000	21:03	
Subject Sample	# 1	21:06	No Sample Provided
Breath Volume	0.000	21:06	
Air Blank	0.000	21:06	
Air Blank	0.000	21:08	
Subject Sample	# 2	21:11	No Sample Provided
Breath Volume	0.000	21:11	
Air Blank	0.000	21:12	
Control Test	0.082	21:12	
Air Blank	0.000	21:13	
Diagnostics Check	OK	21:13	

Data Download Date/Time: 08/17/2018 11:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,J,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: COLLARD

DATE/TIME OF TEST: 07/21/2018 22:26:06

Diagnostics Check	OK	22:23	
Air Blank	0.000	22:23	
Control Test	0.081	22:23	
Air Blank	0.000	22:24	
Subject Sample	# 1	22:24	Subject Test Refused
Breath Volume	0.000	22:24	
Air Blank	0.000	22:25	
Control Test	0.080	22:25	
Air Blank	0.000	22:26	
Diagnostics Check	OK	22:26	

Data Download Date/Time: 08/17/2018 11:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,J,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISSETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: VAUGHN

DATE/TIME OF TEST: 07/21/2018 22:56:42

Diagnostics Check	OK	22:53	
Air Blank	0.000	22:54	
Control Test	0.081	22:54	
Air Blank	0.000	22:55	
Subject Sample	# 1	22:55	Subject Test Refused
Breath Volume	0.000	22:55	
Air Blank	0.000	22:55	
Control Test	0.080	22:56	
Air Blank	0.000	22:56	
Diagnostics Check	OK	22:56	

Data Download Date/Time: 08/17/2018 11:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,J,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: MCCUTCHEN

DATE/TIME OF TEST: 07/23/2018 02:00:53

Diagnostics Check	OK	01:57	
Air Blank	0.000	01:58	
Control Test	0.080	01:58	
Air Blank	0.000	01:59	
Subject Sample	# 1	01:59	Subject Test Refused
Breath Volume	0.000	01:59	
Air Blank	0.000	01:59	
Control Test	0.080	02:00	
Air Blank	0.000	02:00	
Diagnostics Check	OK	02:00	

Data Download Date/Time: 08/17/2018 11:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,J,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ABREU,A
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:12

SUBJECT'S LAST NAME: SHELTON

DATE/TIME OF TEST: 07/23/2018 05:39:06

Diagnostics Check	OK	05:32
Air Blank	0.000	05:32
Control Test	0.080	05:33
Air Blank	0.000	05:33
Subject Sample # 1	0.100	05:34
Breath Volume	1.250	05:34
Air Blank	0.000	05:35
Air Blank	0.000	05:36
Subject Sample # 2	0.100	05:37
Breath Volume	1.148	05:37
Air Blank	0.000	05:38
Control Test	0.080	05:38
Air Blank	0.000	05:39
Diagnostics Check	OK	05:39

Data Download Date/Time: 08/17/2018 11:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,J,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GARCIA,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:34

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 07/24/2018 20:05:54

Diagnostics Check	OK	19:54	
Air Blank	0.000	19:54	
Control Test	0.081	19:55	
Air Blank	0.000	19:55	
Subject Sample	# 1	19:58	No Sample Provided
Breath Volume	0.000	19:58	
Air Blank	0.000	19:59	
Air Blank	0.000	20:01	
Subject Sample	# 2	20:04	No Sample Provided
Breath Volume	0.000	20:04	
Air Blank	0.000	20:04	
Control Test	0.081	20:05	
Air Blank	0.000	20:05	
Diagnostics Check	OK	20:05	

Data Download Date/Time: 08/17/2018 11:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,J,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CAMACHO,K
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: STEINKE

DATE/TIME OF TEST: 07/24/2018 20:55:32

Diagnostics Check	OK	20:49
Air Blank	0.000	20:49
Control Test	0.081	20:49
Air Blank	0.000	20:50
Subject Sample # 1	0.000	20:51
Breath Volume	3.093	20:51
Air Blank	0.000	20:51
Air Blank	0.000	20:53
Subject Sample # 2	0.000	20:54
Breath Volume	2.636	20:54
Air Blank	0.000	20:54
Control Test	0.082	20:54
Air Blank	0.000	20:55
Diagnostics Check	OK	20:55

Data Download Date/Time: 08/17/2018 11:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,J,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISSETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: HIGGINBOTHAM

DATE/TIME OF TEST: 07/25/2018 03:59:19

Diagnostics Check	OK	03:56	
Air Blank	0.000	03:56	
Control Test	0.081	03:57	
Air Blank	0.000	03:57	
Subject Sample	# 1	03:57	Subject Test Refused
Breath Volume	0.000	03:57	
Air Blank	0.000	03:58	
Control Test	0.079	03:58	
Air Blank	0.000	03:59	
Diagnostics Check	OK	03:59	

Data Download Date/Time: 08/17/2018 11:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURRIELL,E,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: SIMONEAU

DATE/TIME OF TEST: 07/25/2018 21:23:35

Diagnostics Check	OK	21:20	
Air Blank	0.000	21:21	
Control Test	0.081	21:21	
Air Blank	0.000	21:21	
Subject Sample	# 1	21:22	Subject Test Refused
Breath Volume	0.000	21:22	
Air Blank	0.000	21:22	
Control Test	0.080	21:22	
Air Blank	0.000	21:23	
Diagnostics Check	OK	21:23	

Data Download Date/Time: 08/17/2018 11:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRYANT,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: COLE

DATE/TIME OF TEST: 07/28/2018 01:02:23

Diagnostics Check	OK	00:59	
Air Blank	0.000	00:59	
Control Test	0.081	01:00	
Air Blank	0.000	01:00	
Subject Sample	# 1	01:00	Subject Test Refused
Breath Volume	0.000	01:00	
Air Blank	0.000	01:01	
Control Test	0.080	01:01	
Air Blank	0.000	01:02	
Diagnostics Check	OK	01:02	

Data Download Date/Time: 08/17/2018 11:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VANHOF, J
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: LARSON

DATE/TIME OF TEST: 07/28/2018 02:50:36

Diagnostics Check	OK	02:47	
Air Blank	0.000	02:48	
Control Test	0.081	02:48	
Air Blank	0.000	02:49	
Subject Sample	# 1	02:49	Subject Test Refused
Breath Volume	0.000	02:49	
Air Blank	0.000	02:49	
Control Test	0.080	02:49	
Air Blank	0.000	02:50	
Diagnostics Check	OK	02:50	

Data Download Date/Time: 08/17/2018 11:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SOSA, C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: MACK

DATE/TIME OF TEST: 07/28/2018 05:27:49

Diagnostics Check	OK	05:21
Air Blank	0.000	05:21
Control Test	0.080	05:21
Air Blank	0.000	05:22
Subject Sample # 1	0.050	05:22
Breath Volume	1.183	05:22
Air Blank	0.000	05:23
Air Blank	0.000	05:25
Subject Sample # 2	0.058	05:26
Breath Volume	1.246	05:26
Air Blank	0.000	05:26
Control Test	0.081	05:27
Air Blank	0.000	05:27
Diagnostics Check	OK	05:27

Data Download Date/Time: 08/17/2018 11:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONG,JAMES,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SPENCER,K
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 10:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:52

SUBJECT'S LAST NAME: MCMILLIAN

DATE/TIME OF TEST: 07/28/2018 13:20:20

Diagnostics Check	OK	13:13
Air Blank	0.000	13:14
Control Test	0.081	13:14
Air Blank	0.000	13:15
Subject Sample # 1	0.165	13:15
Breath Volume	1.757	13:15
Air Blank	0.000	13:16
Air Blank	0.000	13:18
Subject Sample # 2	0.166	13:18
Breath Volume	2.070	13:18
Air Blank	0.000	13:19
Control Test	0.080	13:19
Air Blank	0.000	13:20
Diagnostics Check	OK	13:20

Data Download Date/Time: 08/17/2018 11:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRYANT, R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 07/29/2018 00:49:51

Diagnostics Check	OK	00:42
Air Blank	0.000	00:43
Control Test	0.080	00:43
Air Blank	0.000	00:44
Subject Sample # 1	0.186	00:44
Breath Volume	2.292	00:44
Air Blank	0.000	00:45
Air Blank	0.000	00:47
Subject Sample # 2	0.184	00:48
Breath Volume	1.359	00:48
Air Blank	0.000	00:48
Control Test	0.079	00:49
Air Blank	0.000	00:49
Diagnostics Check	OK	00:49

Data Download Date/Time: 08/17/2018 11:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BROWN, T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: MCQUILLIA

DATE/TIME OF TEST: 07/29/2018 04:09:11

Diagnostics Check	OK	04:02
Air Blank	0.000	04:03
Control Test	0.081	04:03
Air Blank	0.000	04:03
Subject Sample # 1	0.130	04:04
Breath Volume	1.140	04:04
Air Blank	0.000	04:05
Air Blank	0.000	04:07
Subject Sample # 2	0.134	04:07
Breath Volume	1.132	04:07
Air Blank	0.000	04:08
Control Test	0.080	04:08
Air Blank	0.000	04:09
Diagnostics Check	OK	04:09

Data Download Date/Time: 08/17/2018 11:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONG,JAMES,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CLARK,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:27

SUBJECT'S LAST NAME: LUCERO

DATE/TIME OF TEST: 07/29/2018 07:53:05

Diagnostics Check	OK	07:49	
Air Blank	0.000	07:50	
Control Test	0.081	07:50	
Air Blank	0.000	07:51	
Subject Sample	# 1	07:51	Subject Test Refused
Breath Volume	0.000	07:51	
Air Blank	0.000	07:52	
Control Test	0.080	07:52	
Air Blank	0.000	07:53	
Diagnostics Check	OK	07:53	

Data Download Date/Time: 08/17/2018 11:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MARRERO, J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 07/30/2018 00:23:35

Diagnostics Check	OK	00:20	
Air Blank	0.000	00:21	
Control Test	0.081	00:21	
Air Blank	0.000	00:21	
Subject Sample	# 1	00:22	Subject Test Refused
Breath Volume	0.000	00:22	
Air Blank	0.000	00:22	
Control Test	0.080	00:22	
Air Blank	0.000	00:23	
Diagnostics Check	OK	00:23	

Data Download Date/Time: 08/17/2018 11:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BROWN, T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: RAMOS ULLOA

DATE/TIME OF TEST: 07/30/2018 03:14:25

Diagnostics Check	OK	03:07
Air Blank	0.000	03:07
Control Test	0.081	03:07
Air Blank	0.000	03:08
Subject Sample # 1	0.189	03:09
Breath Volume	1.816	03:09
Air Blank	0.000	03:10
Air Blank	0.000	03:11
Subject Sample # 2	0.187	03:12
Breath Volume	2.191	03:12
Air Blank	0.000	03:13
Control Test	0.080	03:13
Air Blank	0.000	03:14
Diagnostics Check	OK	03:14

Data Download Date/Time: 08/17/2018 11:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BROWN, T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 07/31/2018 02:28:54

Diagnostics Check	OK	02:22
Air Blank	0.000	02:22
Control Test	0.081	02:22
Air Blank	0.000	02:23
Subject Sample # 1	0.123	02:24
Breath Volume	3.878	02:24
Air Blank	0.000	02:24
Air Blank	0.000	02:26
Subject Sample # 2	0.134	02:27
Breath Volume	3.671	02:27
Air Blank	0.000	02:27
Control Test	0.080	02:28
Air Blank	0.000	02:28
Diagnostics Check	OK	02:28

Data Download Date/Time: 08/17/2018 11:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JENSEN, D
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: TRAN

DATE/TIME OF TEST: 07/31/2018 04:29:36

Diagnostics Check	OK	04:22
Air Blank	0.000	04:23
Control Test	0.080	04:23
Air Blank	0.000	04:24
Subject Sample # 1	0.114	04:24
Breath Volume	2.523	04:24
Air Blank	0.000	04:25
Air Blank	0.000	04:27
Subject Sample # 2	0.109	04:27
Breath Volume	2.398	04:27
Air Blank	0.000	04:28
Control Test	0.080	04:28
Air Blank	0.000	04:29
Diagnostics Check	OK	04:29

Data Download Date/Time: 08/17/2018 11:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JASIUKIEWICZ,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: PETERS

DATE/TIME OF TEST: 07/07/2018 03:05:32

Diagnostics Check	OK	03:02	
Air Blank	0.000	03:02	
Control Test	0.079	03:03	
Air Blank	0.000	03:03	
Subject Sample	# 1	03:04	Subject Test Refused
Breath Volume	0.000	03:04	
Air Blank	0.000	03:04	
Control Test	0.079	03:04	
Air Blank	0.000	03:05	
Diagnostics Check	OK	03:05	

Data Download Date/Time: 07/13/2018 12:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOYLE, THOMAS, E
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: MASTERS-SAMRETH, DAVID
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: JORDAN

DATE/TIME OF TEST: 07/07/2018 04:01:33

Diagnostics Check	OK	03:50	
Air Blank	0.000	03:50	
Control Test	0.078	03:50	
Air Blank	0.000	03:51	
Subject Sample	# 1	03:54	No Sample Provided
Breath Volume	0.000	03:54	
Air Blank	0.000	03:55	
Air Blank	0.000	03:57	
Subject Sample	# 2	04:00	No Sample Provided
Breath Volume	0.000	04:00	
Air Blank	0.000	04:00	
Control Test	0.078	04:00	
Air Blank	0.000	04:01	
Diagnostics Check	OK	04:01	

Data Download Date/Time: 07/13/2018 12:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JASIUKIEWICZ,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:28

SUBJECT'S LAST NAME: BLOOM

DATE/TIME OF TEST: 07/07/2018 21:54:14

Diagnostics Check	OK	21:48
Air Blank	0.000	21:48
Control Test	0.079	21:49
Air Blank	0.000	21:49
Subject Sample # 1	0.000	21:50
Breath Volume	1.351	21:50
Air Blank	0.000	21:50
Air Blank	0.000	21:52
Subject Sample # 2	0.000	21:52
Breath Volume	2.031	21:52
Air Blank	0.000	21:53
Control Test	0.081	21:53
Air Blank	0.000	21:54
Diagnostics Check	OK	21:54

Data Download Date/Time: 07/13/2018 12:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JASIUKIEWICZ,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: MITCHEM

DATE/TIME OF TEST: 07/13/2018 03:08:34

Diagnostics Check	OK	03:05	
Air Blank	0.000	03:06	
Control Test	0.078	03:06	
Air Blank	0.000	03:06	
Subject Sample	# 1	03:07	Subject Test Refused
Breath Volume	0.000	03:07	
Air Blank	0.000	03:07	
Control Test	0.079	03:07	
Air Blank	0.000	03:08	
Diagnostics Check	OK	03:08	

Data Download Date/Time: 07/13/2018 12:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 07/13/2018 22:10:09

Diagnostics Check	OK	22:02
Air Blank	0.000	22:02
Control Test	0.077	22:03
Air Blank	0.000	22:03
Subject Sample # 1	0.210	22:05
Breath Volume	1.519	22:05
Air Blank	0.000	22:05
Air Blank	0.000	22:07
Subject Sample # 2	0.197	22:08
Breath Volume	1.812	22:08
Air Blank	0.000	22:09
Control Test	0.080	22:09
Air Blank	0.000	22:10
Diagnostics Check	OK	22:10

Data Download Date/Time: 08/08/2018 09:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,\
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JASIUKIEWICZ,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: SPEELER

DATE/TIME OF TEST: 07/14/2018 02:01:59

Diagnostics Check	OK	01:59	
Air Blank	0.000	01:59	
Control Test	0.078	01:59	
Air Blank	0.000	02:00	
Subject Sample	# 1	02:00	Subject Test Refused
Breath Volume	0.000	02:00	
Air Blank	0.000	02:01	
Control Test	0.079	02:01	
Air Blank	0.000	02:01	
Diagnostics Check	OK	02:01	

Data Download Date/Time: 08/08/2018 09:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEA,JOHN,D
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SHEA,JOHN
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: GLEASON

DATE/TIME OF TEST: 07/14/2018 03:00:14

Diagnostics Check	OK	02:52
Air Blank	0.000	02:53
Control Test	0.079	02:53
Air Blank	0.000	02:54
Subject Sample # 1	0.073	02:55
Breath Volume	2.371	02:55
Air Blank	0.000	02:55
Air Blank	0.000	02:57
Subject Sample # 2	0.074	02:58
Breath Volume	2.078	02:58
Air Blank	0.000	02:59
Control Test	0.081	02:59
Air Blank	0.000	03:00
Diagnostics Check	OK	03:00

Data Download Date/Time: 08/08/2018 09:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: BUENO DE CAMARGO

DATE/TIME OF TEST: 07/14/2018 03:43:59

Diagnostics Check	OK	03:40	
Air Blank	0.000	03:41	
Control Test	0.079	03:41	
Air Blank	0.000	03:42	
Subject Sample	# 1	03:42	Subject Test Refused
Breath Volume	0.000	03:42	
Air Blank	0.000	03:43	
Control Test	0.079	03:43	
Air Blank	0.000	03:43	
Diagnostics Check	OK	03:43	

Data Download Date/Time: 08/08/2018 09:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEA,JOHN,D
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SHEA,JOHN
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:52

SUBJECT'S LAST NAME: SWITZER

DATE/TIME OF TEST: 07/16/2018 17:19:27

Diagnostics Check	OK	17:13
Air Blank	0.000	17:13
Control Test	0.077	17:14
Air Blank	0.000	17:14
Subject Sample # 1	0.302	17:15
Breath Volume	3.054	17:15
Air Blank	0.000	17:15
Air Blank	0.000	17:17
Subject Sample # 2	0.303	17:17
Breath Volume	2.070	17:17
Air Blank	0.000	17:18
Control Test	0.081	17:18
Air Blank	0.000	17:19
Diagnostics Check	OK	17:19

Data Download Date/Time: 08/08/2018 09:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOLBICKA,LAWRENCE,R
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: KOLBICKA,LAWRENCE
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 09:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:31

SUBJECT'S LAST NAME: PETERS

DATE/TIME OF TEST: 07/18/2018 09:57:30

Diagnostics Check	OK	09:51
Air Blank	0.000	09:51
Control Test	0.079	09:52
Air Blank	0.000	09:52
Subject Sample # 1	0.000	09:53
Breath Volume	1.679	09:53
Air Blank	0.000	09:53
Air Blank	0.000	09:55
Subject Sample # 2	0.000	09:56
Breath Volume	1.746	09:56
Air Blank	0.000	09:56
Control Test	0.081	09:56
Air Blank	0.000	09:57
Diagnostics Check	OK	09:57

Data Download Date/Time: 08/08/2018 09:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOYLE,THOMAS,E
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: BOYLE,THOMAS
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:55

SUBJECT'S LAST NAME: DUEHRING

DATE/TIME OF TEST: 07/20/2018 18:19:01

Diagnostics Check	OK	18:15	
Air Blank	0.000	18:16	
Control Test	0.077	18:16	
Air Blank	0.000	18:17	
Subject Sample	# 1	18:17	Slope Not Met
Breath Volume	1.007	18:17	
Air Blank	0.000	18:18	
Control Test	0.079	18:18	
Air Blank	0.000	18:18	
Diagnostics Check	OK	18:19	

Data Download Date/Time: 08/08/2018 09:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOYLE,THOMAS,E
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: BOYLE,THOMAS
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: DUEHRING

DATE/TIME OF TEST: 07/20/2018 18:46:13

Diagnostics Check	OK	18:40
Air Blank	0.000	18:40
Control Test	0.078	18:40
Air Blank	0.000	18:41
Subject Sample # 1	0.208	18:41
Breath Volume	1.824	18:41
Air Blank	0.000	18:42
Air Blank	0.000	18:44
Subject Sample # 2	0.207	18:44
Breath Volume	1.843	18:44
Air Blank	0.000	18:45
Control Test	0.079	18:45
Air Blank	0.000	18:46
Diagnostics Check	OK	18:46

Data Download Date/Time: 08/08/2018 09:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSEN,STEPHEN,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: HOLM,MATTHEW
ARRESTING OFFICER AGENCY: Indian Shores PD
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: WINGFIELD

DATE/TIME OF TEST: 07/20/2018 23:26:06

Diagnostics Check	OK	23:23	
Air Blank	0.000	23:23	
Control Test	0.078	23:24	
Air Blank	0.000	23:24	
Subject Sample	# 1	23:24	Subject Test Refused
Breath Volume	0.000	23:24	
Air Blank	0.000	23:25	
Control Test	0.077	23:25	
Air Blank	0.000	23:26	
Diagnostics Check	OK	23:26	

Data Download Date/Time: 08/08/2018 09:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOYLE,THOMAS,E
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: BOYLE,THOMAS
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: MCCOWN

DATE/TIME OF TEST: 07/20/2018 23:47:22

Diagnostics Check	OK	23:38	
Air Blank	0.000	23:39	
Control Test	0.078	23:39	
Air Blank	0.000	23:40	
Subject Sample	# 1	23:43	No Sample Provided
Breath Volume	0.000	23:43	
Air Blank	0.000	23:43	
Air Blank	0.000	23:45	
Subject Sample	# 2	23:46	Subject Test Refused
Breath Volume	0.000	23:46	
Air Blank	0.000	23:46	
Control Test	0.078	23:46	
Air Blank	0.000	23:47	
Diagnostics Check	OK	23:47	

Data Download Date/Time: 08/08/2018 09:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOYLE,THOMAS,E
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: BOYLE,THOMAS
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: ZEVAKOV

DATE/TIME OF TEST: 07/22/2018 03:07:38

Diagnostics Check	OK	03:04	
Air Blank	0.000	03:05	
Control Test	0.079	03:05	
Air Blank	0.000	03:06	
Subject Sample	# 1	03:06	Subject Test Refused
Breath Volume	0.000	03:06	
Air Blank	0.000	03:06	
Control Test	0.080	03:07	
Air Blank	0.000	03:07	
Diagnostics Check	OK	03:07	

Data Download Date/Time: 08/08/2018 09:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JASIUKIEWICZ,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:28

SUBJECT'S LAST NAME: SAVKO

DATE/TIME OF TEST: 07/25/2018 21:59:37

Diagnostics Check	OK	21:52
Air Blank	0.000	21:52
Control Test	0.078	21:53
Air Blank	0.000	21:53
Subject Sample # 1	0.000	21:55
Breath Volume	2.429	21:55
Air Blank	0.000	21:55
Air Blank	0.000	21:57
Subject Sample # 2	0.000	21:58
Breath Volume	1.539	21:58
Air Blank	0.000	21:58
Control Test	0.081	21:59
Air Blank	0.000	21:59
Diagnostics Check	OK	21:59

Data Download Date/Time: 08/08/2018 09:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ, MICHAEL, P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JASIUKIEWICZ, MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: THALHEIMER

DATE/TIME OF TEST: 07/28/2018 03:48:42

Diagnostics Check	OK	03:42	
Air Blank	0.000	03:42	
Control Test	0.078	03:42	
Air Blank	0.000	03:43	
Subject Sample # 1	0.104	03:43	
Breath Volume	1.160	03:43	
Air Blank	0.000	03:44	
Air Blank	0.000	03:46	
Subject Sample # 2	0.106	03:47	
Breath Volume	1.769	03:47	
Air Blank	0.000	03:47	
Control Test	0.084	03:48	Interferent Detect
Air Blank	0.000	03:48	

Data Download Date/Time: 08/08/2018 09:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JASIUKIEWICZ,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: THALHEIMER

DATE/TIME OF TEST: 07/28/2018 03:56:46

Diagnostics Check	OK	03:50
Air Blank	0.000	03:50
Control Test	0.078	03:51
Air Blank	0.000	03:51
Subject Sample # 1	0.102	03:52
Breath Volume	2.031	03:52
Air Blank	0.000	03:53
Air Blank	0.000	03:54
Subject Sample # 2	0.099	03:55
Breath Volume	1.843	03:55
Air Blank	0.000	03:55
Control Test	0.080	03:56
Air Blank	0.000	03:56
Diagnostics Check	OK	03:56

Data Download Date/Time: 08/08/2018 09:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLEY,SCOTT,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: GALLEY,SCOTT
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 13:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:21

SUBJECT'S LAST NAME: HUDGINS

DATE/TIME OF TEST: 07/30/2018 13:44:20

Diagnostics Check	OK	13:41	
Air Blank	0.000	13:41	
Control Test	0.078	13:41	
Air Blank	0.000	13:42	
Subject Sample	# 1	13:42	Subject Test Refused
Breath Volume	0.000	13:42	
Air Blank	0.000	13:43	
Control Test	0.079	13:43	
Air Blank	0.000	13:44	
Diagnostics Check	OK	13:44	

Data Download Date/Time: 08/08/2018 09:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,C,L
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SALVADOR,E
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: SANDERS

DATE/TIME OF TEST: 07/05/2018 02:43:04

Diagnostics Check	OK	02:36
Air Blank	0.000	02:36
Control Test	0.079	02:37
Air Blank	0.000	02:37
Subject Sample # 1	0.110	02:38
Breath Volume	1.644	02:38
Air Blank	0.000	02:39
Air Blank	0.000	02:40
Subject Sample # 2	0.109	02:41
Breath Volume	1.457	02:41
Air Blank	0.000	02:42
Control Test	0.079	02:42
Air Blank	0.000	02:42
Diagnostics Check	OK	02:43

Data Download Date/Time: 07/28/2018 21:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,C,
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: WHITTINGTON,B
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:39

SUBJECT'S LAST NAME: VASQUEZ

DATE/TIME OF TEST: 07/23/2018 22:02:33

Diagnostics Check	OK	21:59	
Air Blank	0.000	22:00	
Control Test	0.079	22:00	
Air Blank	0.000	22:00	
Subject Sample	# 1	22:01	Subject Test Refused
Breath Volume	0.000	22:01	
Air Blank	0.000	22:01	
Control Test	0.079	22:01	
Air Blank	0.000	22:02	
Diagnostics Check	OK	22:02	

Data Download Date/Time: 07/28/2018 21:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: WALCUTT

DATE/TIME OF TEST: 07/12/2018 01:03:56

Diagnostics Check	OK	00:57
Air Blank	0.000	00:58
Control Test	0.081	00:58
Air Blank	0.000	00:58
Subject Sample # 1	0.091	00:59
Breath Volume	2.855	00:59
Air Blank	0.000	01:00
Air Blank	0.000	01:02
Subject Sample # 2	0.087	01:02
Breath Volume	2.542	01:02
Air Blank	0.000	01:03
Control Test	0.081	01:03
Air Blank	0.000	01:03
Diagnostics Check	OK	01:03

Data Download Date/Time: 08/14/2018 10:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: HO

DATE/TIME OF TEST: 07/15/2018 03:35:05

Diagnostics Check	OK	03:27
Air Blank	0.000	03:27
Control Test	0.079	03:28
Air Blank	0.000	03:28
Subject Sample # 1	0.236	03:30
Breath Volume	1.160	03:30
Air Blank	0.000	03:30
Air Blank	0.000	03:32
Subject Sample # 2	0.237	03:33
Breath Volume	1.210	03:33
Air Blank	0.000	03:34
Control Test	0.079	03:34
Air Blank	0.000	03:35
Diagnostics Check	OK	03:35

Data Download Date/Time: 08/14/2018 10:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: SALAT

DATE/TIME OF TEST: 07/17/2018 01:52:44

Diagnostics Check	OK	01:41	
Air Blank	0.000	01:41	
Control Test	0.079	01:42	
Air Blank	0.000	01:42	
Subject Sample	# 1 0.158	01:45	Volume Not Met
Breath Volume	1.046	01:45	
Air Blank	0.000	01:46	
Air Blank	0.000	01:48	
Subject Sample	# 2	01:51	Subject Test Refused
Breath Volume	0.152	01:51	
Air Blank	0.000	01:51	
Control Test	0.079	01:52	
Air Blank	0.000	01:52	
Diagnostics Check	OK	01:52	

Data Download Date/Time: 08/14/2018 10:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,GAVIN,O
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: HENRY,GAVIN
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: CEDENO

DATE/TIME OF TEST: 07/20/2018 03:39:55

Diagnostics Check	OK	03:33
Air Blank	0.000	03:34
Control Test	0.079	03:34
Air Blank	0.000	03:34
Subject Sample # 1	0.113	03:35
Breath Volume	1.816	03:35
Air Blank	0.000	03:35
Air Blank	0.000	03:37
Subject Sample # 2	0.120	03:38
Breath Volume	1.824	03:38
Air Blank	0.000	03:39
Control Test	0.080	03:39
Air Blank	0.000	03:39
Diagnostics Check	OK	03:39

Data Download Date/Time: 08/14/2018 10:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: MCCORMICK

DATE/TIME OF TEST: 07/24/2018 22:49:14

Diagnostics Check	OK	22:46	
Air Blank	0.000	22:46	
Control Test	0.080	22:47	
Air Blank	0.000	22:47	
Subject Sample	# 1	22:47	Subject Test Refused
Breath Volume	0.000	22:47	
Air Blank	0.000	22:48	
Control Test	0.080	22:48	
Air Blank	0.000	22:49	
Diagnostics Check	OK	22:49	

Data Download Date/Time: 08/14/2018 10:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: COURTNEY

DATE/TIME OF TEST: 07/31/2018 00:56:42

Diagnostics Check	OK	00:53	
Air Blank	0.000	00:53	
Control Test	0.080	00:54	
Air Blank	0.000	00:54	
Subject Sample	# 1	00:55	Subject Test Refused
Breath Volume	0.000	00:55	
Air Blank	0.000	00:55	
Control Test	0.081	00:56	
Air Blank	0.000	00:56	
Diagnostics Check	OK	00:56	

Data Download Date/Time: 08/14/2018 10:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL, EDWARD, A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE, J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: COOP

DATE/TIME OF TEST: 07/01/2018 01:19:19

Diagnostics Check	OK	01:13
Air Blank	0.000	01:13
Control Test	0.079	01:13
Air Blank	0.000	01:14
Subject Sample # 1	0.199	01:14
Breath Volume	1.871	01:14
Air Blank	0.000	01:15
Air Blank	0.000	01:17
Subject Sample # 2	0.194	01:17
Breath Volume	2.070	01:17
Air Blank	0.000	01:18
Control Test	0.078	01:18
Air Blank	0.000	01:19
Diagnostics Check	OK	01:19

Data Download Date/Time: 07/06/2018 12:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: MARKANT

DATE/TIME OF TEST: 07/01/2018 01:57:42

Diagnostics Check	OK	01:54	
Air Blank	0.000	01:55	
Control Test	0.079	01:55	
Air Blank	0.000	01:56	
Subject Sample	# 1	01:56	Subject Test Refused
Breath Volume	0.000	01:56	
Air Blank	0.000	01:56	
Control Test	0.079	01:57	
Air Blank	0.000	01:57	
Diagnostics Check	OK	01:57	

Data Download Date/Time: 07/06/2018 12:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SPLATT,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:04

SUBJECT'S LAST NAME: PAZ

DATE/TIME OF TEST: 07/01/2018 05:31:04

Diagnostics Check	OK	05:24
Air Blank	0.000	05:25
Control Test	0.079	05:25
Air Blank	0.000	05:26
Subject Sample # 1	0.124	05:26
Breath Volume	3.164	05:26
Air Blank	0.000	05:27
Air Blank	0.000	05:29
Subject Sample # 2	0.119	05:29
Breath Volume	3.085	05:29
Air Blank	0.000	05:30
Control Test	0.079	05:30
Air Blank	0.000	05:30
Diagnostics Check	OK	05:31

Data Download Date/Time: 07/06/2018 12:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHRISTMAS,F
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 06/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: BUI
ARREST TIME: 17:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:27

SUBJECT'S LAST NAME: MAHON

DATE/TIME OF TEST: 07/01/2018 18:54:51

Diagnostics Check	OK	18:48
Air Blank	0.000	18:48
Control Test	0.079	18:48
Air Blank	0.000	18:49
Subject Sample # 1	0.089	18:50
Breath Volume	3.335	18:50
Air Blank	0.000	18:50
Air Blank	0.000	18:52
Subject Sample # 2	0.089	18:53
Breath Volume	3.714	18:53
Air Blank	0.000	18:53
Control Test	0.079	18:54
Air Blank	0.000	18:54
Diagnostics Check	OK	18:54

Data Download Date/Time: 07/06/2018 12:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRYANT,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: TOWNS

DATE/TIME OF TEST: 07/02/2018 00:03:45

Diagnostics Check	OK	00:00	
Air Blank	0.000	00:01	
Control Test	0.079	00:01	
Air Blank	0.000	00:02	
Subject Sample	# 1	00:02	Subject Test Refused
Breath Volume	0.000	00:02	
Air Blank	0.000	00:02	
Control Test	0.079	00:03	
Air Blank	0.000	00:03	
Diagnostics Check	OK	00:03	

Data Download Date/Time: 07/06/2018 12:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BROWN,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:11

SUBJECT'S LAST NAME: TUCKER

DATE/TIME OF TEST: 07/02/2018 21:38:46

Diagnostics Check	OK	21:32
Air Blank	0.000	21:32
Control Test	0.079	21:33
Air Blank	0.000	21:33
Subject Sample # 1	0.000	21:34
Breath Volume	3.570	21:34
Air Blank	0.000	21:34
Air Blank	0.000	21:36
Subject Sample # 2	0.000	21:37
Breath Volume	3.195	21:37
Air Blank	0.000	21:37
Control Test	0.078	21:38
Air Blank	0.000	21:38
Diagnostics Check	OK	21:38

Data Download Date/Time: 07/06/2018 12:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MARRERO,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 07/03/2018 02:51:02

Diagnostics Check	OK	02:44
Air Blank	0.000	02:45
Control Test	0.079	02:45
Air Blank	0.000	02:46
Subject Sample # 1	0.171	02:46
Breath Volume	3.687	02:46
Air Blank	0.000	02:47
Air Blank	0.000	02:49
Subject Sample # 2	0.171	02:49
Breath Volume	3.648	02:49
Air Blank	0.000	02:50
Control Test	0.079	02:50
Air Blank	0.000	02:50
Diagnostics Check	OK	02:51

Data Download Date/Time: 07/06/2018 12:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ZURN,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: ROWLAND

DATE/TIME OF TEST: 07/03/2018 23:22:55

Diagnostics Check	OK	23:16
Air Blank	0.000	23:16
Control Test	0.079	23:16
Air Blank	0.000	23:17
Subject Sample # 1	0.197	23:17
Breath Volume	2.148	23:17
Air Blank	0.000	23:18
Air Blank	0.000	23:20
Subject Sample # 2	0.199	23:21
Breath Volume	2.191	23:21
Air Blank	0.000	23:21
Control Test	0.078	23:22
Air Blank	0.000	23:22
Diagnostics Check	OK	23:22

Data Download Date/Time: 07/06/2018 12:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: KELSO

DATE/TIME OF TEST: 07/04/2018 02:20:46

Diagnostics Check	OK	02:14
Air Blank	0.000	02:14
Control Test	0.079	02:15
Air Blank	0.000	02:15
Subject Sample # 1	0.183	02:16
Breath Volume	3.496	02:16
Air Blank	0.000	02:16
Air Blank	0.000	02:18
Subject Sample # 2	0.192	02:19
Breath Volume	2.730	02:19
Air Blank	0.000	02:19
Control Test	0.078	02:20
Air Blank	0.000	02:20
Diagnostics Check	OK	02:20

Data Download Date/Time: 07/06/2018 12:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KASSIS,F
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: LIVINGSTON

DATE/TIME OF TEST: 07/04/2018 23:23:23

Diagnostics Check	OK	23:20	
Air Blank	0.000	23:20	
Control Test	0.079	23:21	
Air Blank	0.000	23:21	
Subject Sample	# 1	23:21	Subject Test Refused
Breath Volume	0.000	23:21	
Air Blank	0.000	23:22	
Control Test	0.079	23:22	
Air Blank	0.000	23:23	
Diagnostics Check	OK	23:23	

Data Download Date/Time: 07/06/2018 12:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: KOESTENS

DATE/TIME OF TEST: 07/05/2018 03:19:42

Diagnostics Check	OK	03:13
Air Blank	0.000	03:14
Control Test	0.080	03:14
Air Blank	0.000	03:15
Subject Sample # 1	0.026	03:15
Breath Volume	1.859	03:15
Air Blank	0.000	03:15
Air Blank	0.000	03:17
Subject Sample # 2	0.024	03:18
Breath Volume	1.511	03:18
Air Blank	0.000	03:18
Control Test	0.079	03:19
Air Blank	0.000	03:19
Diagnostics Check	OK	03:19

Data Download Date/Time: 07/06/2018 12:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: RAY

DATE/TIME OF TEST: 07/05/2018 04:13:05

Diagnostics Check	OK	04:10	
Air Blank	0.000	04:10	
Control Test	0.079	04:10	
Air Blank	0.000	04:11	
Subject Sample	# 1	04:11	Subject Test Refused
Breath Volume	0.000	04:11	
Air Blank	0.000	04:12	
Control Test	0.079	04:12	
Air Blank	0.000	04:12	
Diagnostics Check	OK	04:13	

Data Download Date/Time: 07/06/2018 12:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL, EDWARD, A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE, J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: LUCAS

DATE/TIME OF TEST: 07/06/2018 01:41:35

Diagnostics Check	OK	01:35
Air Blank	0.000	01:36
Control Test	0.079	01:36
Air Blank	0.000	01:36
Subject Sample # 1	0.000	01:37
Breath Volume	3.816	01:37
Air Blank	0.000	01:37
Air Blank	0.000	01:39
Subject Sample # 2	0.000	01:40
Breath Volume	3.878	01:40
Air Blank	0.000	01:40
Control Test	0.079	01:40
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

Data Download Date/Time: 07/06/2018 12:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MORRISON,M
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 07/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:57

SUBJECT'S LAST NAME: LAMBERT

DATE/TIME OF TEST: 07/06/2018 05:23:34

Diagnostics Check	OK	05:17
Air Blank	0.000	05:17
Control Test	0.080	05:18
Air Blank	0.000	05:18
Subject Sample # 1	0.175	05:19
Breath Volume	3.097	05:19
Air Blank	0.000	05:19
Air Blank	0.000	05:21
Subject Sample # 2	0.172	05:21
Breath Volume	2.359	05:21
Air Blank	0.000	05:22
Control Test	0.079	05:22
Air Blank	0.000	05:23
Diagnostics Check	OK	05:23

Data Download Date/Time: 07/06/2018 12:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILBERGER,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:16

SUBJECT'S LAST NAME: POTYRAJ

DATE/TIME OF TEST: 07/07/2018 20:43:26

Diagnostics Check	OK	20:36
Air Blank	0.000	20:37
Control Test	0.079	20:37
Air Blank	0.000	20:38
Subject Sample # 1	0.214	20:38
Breath Volume	2.511	20:38
Air Blank	0.000	20:39
Air Blank	0.000	20:41
Subject Sample # 2	0.208	20:41
Breath Volume	2.617	20:41
Air Blank	0.000	20:42
Control Test	0.079	20:42
Air Blank	0.000	20:43
Diagnostics Check	OK	20:43

Data Download Date/Time: 07/18/2018 14:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILBERGER,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BROWN,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: CROCKER JR

DATE/TIME OF TEST: 07/07/2018 22:25:32

Diagnostics Check	OK	22:22	
Air Blank	0.000	22:23	
Control Test	0.079	22:23	
Air Blank	0.000	22:23	
Subject Sample	# 1	22:24	Subject Test Refused
Breath Volume	0.000	22:24	
Air Blank	0.000	22:24	
Control Test	0.079	22:24	
Air Blank	0.000	22:25	
Diagnostics Check	OK	22:25	

Data Download Date/Time: 07/18/2018 14:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAGE,D,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FOWLER,W
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 13:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:13

SUBJECT'S LAST NAME: OCONNOR

DATE/TIME OF TEST: 07/08/2018 15:50:36

Diagnostics Check	OK	15:44
Air Blank	0.000	15:44
Control Test	0.080	15:45
Air Blank	0.000	15:45
Subject Sample # 1	0.000	15:46
Breath Volume	2.042	15:46
Air Blank	0.000	15:46
Air Blank	0.000	15:48
Subject Sample # 2	0.000	15:49
Breath Volume	2.343	15:49
Air Blank	0.000	15:49
Control Test	0.080	15:49
Air Blank	0.000	15:50
Diagnostics Check	OK	15:50

Data Download Date/Time: 07/18/2018 14:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAGE,D,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PETERS,R
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 07/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:11

SUBJECT'S LAST NAME: FREEMAN

DATE/TIME OF TEST: 07/09/2018 19:38:07

Diagnostics Check	OK	19:31
Air Blank	0.000	19:32
Control Test	0.079	19:32
Air Blank	0.000	19:32
Subject Sample # 1	0.184	19:33
Breath Volume	1.808	19:33
Air Blank	0.000	19:34
Air Blank	0.000	19:35
Subject Sample # 2	0.191	19:36
Breath Volume	2.214	19:36
Air Blank	0.000	19:37
Control Test	0.079	19:37
Air Blank	0.000	19:38
Diagnostics Check	OK	19:38

Data Download Date/Time: 07/18/2018 14:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLDEN,E,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 07/10/2018 17:47:11

Diagnostics Check	OK	17:40
Air Blank	0.000	17:41
Control Test	0.079	17:41
Air Blank	0.000	17:41
Subject Sample # 1	0.212	17:42
Breath Volume	1.210	17:42
Air Blank	0.000	17:42
Air Blank	0.000	17:44
Subject Sample # 2	0.198	17:45
Breath Volume	1.472	17:45
Air Blank	0.000	17:46
Control Test	0.078	17:46
Air Blank	0.000	17:47
Diagnostics Check	OK	17:47

Data Download Date/Time: 07/18/2018 14:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAGE,D,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRYANT,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: BAXLEY

DATE/TIME OF TEST: 07/10/2018 22:46:28

Diagnostics Check	OK	22:40
Air Blank	0.000	22:40
Control Test	0.079	22:40
Air Blank	0.000	22:41
Subject Sample # 1	0.139	22:41
Breath Volume	2.449	22:41
Air Blank	0.000	22:42
Air Blank	0.000	22:44
Subject Sample # 2	0.136	22:44
Breath Volume	3.238	22:44
Air Blank	0.000	22:45
Control Test	0.078	22:45
Air Blank	0.000	22:46
Diagnostics Check	OK	22:46

Data Download Date/Time: 07/18/2018 14:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 07/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: BRAMER

DATE/TIME OF TEST: 07/11/2018 03:56:20

Diagnostics Check	OK	03:49
Air Blank	0.000	03:49
Control Test	0.079	03:49
Air Blank	0.000	03:50
Subject Sample # 1	0.170	03:51
Breath Volume	2.011	03:51
Air Blank	0.000	03:52
Air Blank	0.000	03:54
Subject Sample # 2	0.171	03:54
Breath Volume	1.863	03:54
Air Blank	0.000	03:55
Control Test	0.078	03:55
Air Blank	0.000	03:56
Diagnostics Check	OK	03:56

Data Download Date/Time: 07/18/2018 14:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PATTERSON,K
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 07/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:26

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 07/11/2018 19:53:44

Diagnostics Check	OK	19:46
Air Blank	0.000	19:46
Control Test	0.079	19:47
Air Blank	0.000	19:47
Subject Sample # 1	0.193	19:48
Breath Volume	2.152	19:48
Air Blank	0.000	19:49
Air Blank	0.000	19:50
Subject Sample # 2	0.182	19:52
Breath Volume	2.265	19:52
Air Blank	0.000	19:52
Control Test	0.078	19:53
Air Blank	0.000	19:53
Diagnostics Check	OK	19:53

Data Download Date/Time: 07/18/2018 14:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,V,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CAMACHO,K
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 07/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:33

SUBJECT'S LAST NAME: SHOCKLEY

DATE/TIME OF TEST: 07/14/2018 20:00:44

Diagnostics Check	OK	19:53
Air Blank	0.000	19:53
Control Test	0.079	19:54
Air Blank	0.000	19:54
Subject Sample # 1	0.306	19:56
Breath Volume	2.152	19:56
Air Blank	0.000	19:56
Air Blank	0.000	19:58
Subject Sample # 2	0.305	19:59
Breath Volume	1.613	19:59
Air Blank	0.000	19:59
Control Test	0.078	20:00
Air Blank	0.000	20:00
Diagnostics Check	OK	20:00

Data Download Date/Time: 07/18/2018 14:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AUGUSTINE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: WADDELL

DATE/TIME OF TEST: 07/20/2018 02:42:55

Diagnostics Check	OK	02:40	
Air Blank	0.000	02:40	
Control Test	0.080	02:40	
Air Blank	0.000	02:41	
Subject Sample	# 1	02:41	Subject Test Refused
Breath Volume	0.000	02:41	
Air Blank	0.000	02:41	
Control Test	0.080	02:42	
Air Blank	0.000	02:42	
Diagnostics Check	OK	02:42	

Data Download Date/Time: 08/17/2018 11:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AUGUSTINE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 07/20/2018 03:57:24

Diagnostics Check	OK	03:50
Air Blank	0.000	03:50
Control Test	0.080	03:51
Air Blank	0.000	03:51
Subject Sample # 1	0.130	03:52
Breath Volume	2.574	03:52
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample # 2	0.131	03:55
Breath Volume	2.527	03:55
Air Blank	0.000	03:56
Control Test	0.079	03:56
Air Blank	0.000	03:57
Diagnostics Check	OK	03:57

Data Download Date/Time: 08/17/2018 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARIDA,K,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FRANZ,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 06:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:57

SUBJECT'S LAST NAME: LAWSON

DATE/TIME OF TEST: 07/22/2018 11:24:40

Diagnostics Check	OK	11:18
Air Blank	0.000	11:18
Control Test	0.080	11:18
Air Blank	0.000	11:19
Subject Sample # 1	0.053	11:20
Breath Volume	1.914	11:20
Air Blank	0.000	11:20
Air Blank	0.000	11:22
Subject Sample # 2	0.053	11:23
Breath Volume	2.335	11:23
Air Blank	0.000	11:23
Control Test	0.080	11:24
Air Blank	0.000	11:24
Diagnostics Check	OK	11:24

Data Download Date/Time: 08/17/2018 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AUGUSTINE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: SPAULDING

DATE/TIME OF TEST: 07/26/2018 04:45:26

Diagnostics Check	OK	04:39
Air Blank	0.000	04:39
Control Test	0.080	04:39
Air Blank	0.000	04:40
Subject Sample # 1	0.111	04:40
Breath Volume	1.839	04:40
Air Blank	0.000	04:41
Air Blank	0.000	04:43
Subject Sample # 2	0.108	04:43
Breath Volume	1.859	04:43
Air Blank	0.000	04:44
Control Test	0.079	04:44
Air Blank	0.000	04:45
Diagnostics Check	OK	04:45

Data Download Date/Time: 08/17/2018 11:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HARPER,T
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:09

SUBJECT'S LAST NAME: RILEY

DATE/TIME OF TEST: 07/28/2018 05:34:08

Diagnostics Check	OK	05:31	
Air Blank	0.000	05:31	
Control Test	0.080	05:31	
Air Blank	0.000	05:32	
Subject Sample	# 1	05:32	Subject Test Refused
Breath Volume	0.000	05:32	
Air Blank	0.000	05:33	
Control Test	0.079	05:33	
Air Blank	0.000	05:34	
Diagnostics Check	OK	05:34	

Data Download Date/Time: 08/17/2018 11:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: SAM ZA

DATE/TIME OF TEST: 07/28/2018 23:25:20

Diagnostics Check	OK	23:19
Air Blank	0.000	23:19
Control Test	0.080	23:20
Air Blank	0.000	23:20
Subject Sample # 1	0.027	23:20
Breath Volume	1.800	23:20
Air Blank	0.000	23:21
Air Blank	0.000	23:23
Subject Sample # 2	0.028	23:23
Breath Volume	1.453	23:23
Air Blank	0.000	23:24
Control Test	0.080	23:24
Air Blank	0.000	23:25
Diagnostics Check	OK	23:25

Data Download Date/Time: 08/17/2018 11:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MARRERO,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: JAMES

DATE/TIME OF TEST: 07/29/2018 03:00:21

Diagnostics Check	OK	02:53
Air Blank	0.000	02:54
Control Test	0.079	02:54
Air Blank	0.000	02:55
Subject Sample # 1	0.151	02:55
Breath Volume	2.597	02:55
Air Blank	0.000	02:56
Air Blank	0.000	02:58
Subject Sample # 2	0.164	02:58
Breath Volume	1.855	02:58
Air Blank	0.000	02:59
Control Test	0.079	02:59
Air Blank	0.000	03:00
Diagnostics Check	OK	03:00

Data Download Date/Time: 08/17/2018 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: COVINGTON

DATE/TIME OF TEST: 07/29/2018 04:39:26

Diagnostics Check	OK	04:36	
Air Blank	0.000	04:37	
Control Test	0.079	04:37	
Air Blank	0.000	04:37	
Subject Sample	# 1	04:38	Subject Test Refused
Breath Volume	0.000	04:38	
Air Blank	0.000	04:38	
Control Test	0.080	04:38	
Air Blank	0.000	04:39	
Diagnostics Check	OK	04:39	

Data Download Date/Time: 08/17/2018 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARIDA,K,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PETERS,R
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:12

SUBJECT'S LAST NAME: KHAREL

DATE/TIME OF TEST: 07/29/2018 19:39:28

Diagnostics Check	OK	19:33
Air Blank	0.000	19:33
Control Test	0.080	19:33
Air Blank	0.000	19:34
Subject Sample # 1	0.121	19:34
Breath Volume	1.832	19:34
Air Blank	0.000	19:35
Air Blank	0.000	19:37
Subject Sample # 2	0.122	19:37
Breath Volume	1.902	19:37
Air Blank	0.000	19:38
Control Test	0.079	19:38
Air Blank	0.000	19:39
Diagnostics Check	OK	19:39

Data Download Date/Time: 08/17/2018 11:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY, SHANON, E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRYANT, R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: FUDGE

DATE/TIME OF TEST: 07/30/2018 00:48:04

Diagnostics Check	OK	00:41
Air Blank	0.000	00:42
Control Test	0.079	00:42
Air Blank	0.000	00:43
Subject Sample # 1	0.106	00:43
Breath Volume	1.996	00:43
Air Blank	0.000	00:44
Air Blank	0.000	00:45
Subject Sample # 2	0.104	00:46
Breath Volume	1.996	00:46
Air Blank	0.000	00:47
Control Test	0.079	00:47
Air Blank	0.000	00:47
Diagnostics Check	OK	00:48

Data Download Date/Time: 08/17/2018 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY, SHANON, E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRYANT, R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:17

SUBJECT'S LAST NAME: LI

DATE/TIME OF TEST: 07/30/2018 05:44:59

Diagnostics Check	OK	05:38
Air Blank	0.000	05:39
Control Test	0.080	05:39
Air Blank	0.000	05:39
Subject Sample # 1	0.169	05:40
Breath Volume	1.761	05:40
Air Blank	0.000	05:41
Air Blank	0.000	05:43
Subject Sample # 2	0.169	05:43
Breath Volume	1.582	05:43
Air Blank	0.000	05:44
Control Test	0.079	05:44
Air Blank	0.000	05:44
Diagnostics Check	OK	05:44

Data Download Date/Time: 08/17/2018 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY, SHANON, E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE, J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: SANDERS

DATE/TIME OF TEST: 07/31/2018 03:41:37

Diagnostics Check	OK	03:35
Air Blank	0.000	03:35
Control Test	0.080	03:36
Air Blank	0.000	03:36
Subject Sample # 1	0.147	03:37
Breath Volume	2.906	03:37
Air Blank	0.000	03:37
Air Blank	0.000	03:39
Subject Sample # 2	0.154	03:40
Breath Volume	1.796	03:40
Air Blank	0.000	03:40
Control Test	0.079	03:41
Air Blank	0.000	03:41
Diagnostics Check	OK	03:41

Data Download Date/Time: 08/17/2018 11:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNOW,CHRIS,D
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: MILLS,D
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: MANSFIELD

DATE/TIME OF TEST: 07/04/2018 04:57:50

Diagnostics Check	OK	04:46	
Air Blank	0.000	04:46	
Control Test	0.080	04:47	
Air Blank	0.000	04:47	
Subject Sample	# 1 0.054	04:50	Volume Not Met
Breath Volume	1.140	04:50	
Air Blank	0.000	04:51	
Air Blank	0.000	04:53	
Subject Sample	# 2 0.046	04:56	Volume Not Met
Breath Volume	0.519	04:56	
Air Blank	0.000	04:56	
Control Test	0.080	04:57	
Air Blank	0.000	04:57	
Diagnostics Check	OK	04:57	

Data Download Date/Time: 07/25/2018 23:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNOW,CHRIS,D
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: BRYAN,J
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: BUI
ARREST TIME: 19:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: HELLARD JR

DATE/TIME OF TEST: 07/07/2018 20:59:41

Diagnostics Check	OK	20:52
Air Blank	0.000	20:53
Control Test	0.080	20:53
Air Blank	0.000	20:54
Subject Sample # 1	0.128	20:55
Breath Volume	1.617	20:55
Air Blank	0.000	20:55
Air Blank	0.000	20:57
Subject Sample # 2	0.119	20:58
Breath Volume	1.222	20:58
Air Blank	0.000	20:58
Control Test	0.080	20:59
Air Blank	0.000	20:59
Diagnostics Check	OK	20:59

Data Download Date/Time: 07/25/2018 23:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNOW,CHRIS,D
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: MALLEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: DELGADO

DATE/TIME OF TEST: 07/12/2018 23:06:08

Diagnostics Check	OK	22:59
Air Blank	0.000	22:59
Control Test	0.080	23:00
Air Blank	0.000	23:00
Subject Sample # 1	0.130	23:01
Breath Volume	1.546	23:01
Air Blank	0.000	23:01
Air Blank	0.000	23:03
Subject Sample # 2	0.127	23:04
Breath Volume	1.398	23:04
Air Blank	0.000	23:05
Control Test	0.080	23:05
Air Blank	0.000	23:06
Diagnostics Check	OK	23:06

Data Download Date/Time: 07/25/2018 23:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTON,DEPUTY,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: CORDERIO,TROOPER
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 11:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:10

SUBJECT'S LAST NAME: ;6360100716006489448

DATE/TIME OF TEST: 07/21/2018 12:43:10

Diagnostics Check	OK	12:36
Air Blank	0.000	12:37
Control Test	0.080	12:37
Air Blank	0.000	12:38
Subject Sample # 1	0.211	12:38
Breath Volume	2.246	12:38
Air Blank	0.000	12:39
Air Blank	0.000	12:41
Subject Sample # 2	0.213	12:41
Breath Volume	1.433	12:41
Air Blank	0.000	12:42
Control Test	0.080	12:42
Air Blank	0.000	12:43
Diagnostics Check	OK	12:43

Data Download Date/Time: 07/25/2018 23:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,C,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: VEREEN,D
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: POWELL

DATE/TIME OF TEST: 07/22/2018 03:37:04

Diagnostics Check	OK	03:30
Air Blank	0.000	03:31
Control Test	0.080	03:31
Air Blank	0.000	03:32
Subject Sample # 1	0.000	03:32
Breath Volume	3.386	03:32
Air Blank	0.000	03:32
Air Blank	0.000	03:34
Subject Sample # 2	0.000	03:35
Breath Volume	2.968	03:35
Air Blank	0.000	03:36
Control Test	0.080	03:36
Air Blank	0.000	03:36
Diagnostics Check	OK	03:37

Data Download Date/Time: 07/25/2018 23:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTSTATT,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: HENDLEY,R
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 07/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: HANSON

DATE/TIME OF TEST: 07/27/2018 03:34:18

Diagnostics Check	OK	03:25	
Air Blank	0.000	03:26	
Control Test	0.080	03:26	
Air Blank	0.000	03:27	
Subject Sample	# 1	03:30	No Sample Provided
Breath Volume	0.000	03:30	
Air Blank	0.000	03:30	
Air Blank	0.000	03:32	
Subject Sample	# 2	03:32	Subject Test Refused
Breath Volume	0.000	03:32	
Air Blank	0.000	03:33	
Control Test	0.080	03:33	
Air Blank	0.000	03:34	
Diagnostics Check	OK	03:34	

Data Download Date/Time: 08/25/2018 14:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001048

INSTRUMENT REGISTERED TO: LAKE ALFRED PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,LATORIAN,T
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: TURNER,LATORIAN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: WOODDELL

DATE/TIME OF TEST: 07/04/2018 23:37:36

Diagnostics Check	OK	23:28
Air Blank	0.000	23:29
Control Test	0.080	23:29
Air Blank	0.000	23:29
Subject Sample # 1	0.233	23:32
Breath Volume	1.429	23:32
Air Blank	0.000	23:33
Air Blank	0.000	23:34
Subject Sample # 2	0.228	23:35
Breath Volume	1.343	23:35
Air Blank	0.000	23:36
Control Test	0.078	23:36
Air Blank	0.000	23:37
Diagnostics Check	OK	23:37

Data Download Date/Time: 07/19/2018 10:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEDULLA,VINCENT,A
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: PULASKI,D
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A1
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: BUI
ARREST TIME: 18:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: ESENBERG

DATE/TIME OF TEST: 07/01/2018 20:47:22

Diagnostics Check	OK	20:41
Air Blank	0.000	20:41
Control Test	0.079	20:41
Air Blank	0.000	20:42
Subject Sample # 1	0.041	20:42
Breath Volume	4.027	20:42
Air Blank	0.000	20:43
Air Blank	0.000	20:45
Subject Sample # 2	0.040	20:45
Breath Volume	2.726	20:45
Air Blank	0.000	20:46
Control Test	0.080	20:46
Air Blank	0.000	20:47
Diagnostics Check	OK	20:47

Data Download Date/Time: 07/11/2018 14:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARTNER,ARTHUR,F
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GARTNER,ARTHUR
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A1
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: RIVERA MORENO

DATE/TIME OF TEST: 07/12/2018 01:41:58

Diagnostics Check	OK	01:35
Air Blank	0.000	01:36
Control Test	0.080	01:36
Air Blank	0.000	01:37
Subject Sample # 1	0.000	01:37
Breath Volume	2.621	01:37
Air Blank	0.000	01:38
Air Blank	0.000	01:40
Subject Sample # 2	0.000	01:40
Breath Volume	2.722	01:40
Air Blank	0.000	01:41
Control Test	0.079	01:41
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

Data Download Date/Time: 08/03/2018 14:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNORR,PETER,J
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: KNORR,PETER
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A1
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: CATELLIER

DATE/TIME OF TEST: 07/13/2018 21:52:42

Diagnostics Check	OK	21:45
Air Blank	0.000	21:46
Control Test	0.080	21:46
Air Blank	0.000	21:47
Subject Sample # 1	0.109	21:47
Breath Volume	2.675	21:47
Air Blank	0.000	21:48
Air Blank	0.000	21:50
Subject Sample # 2	0.108	21:51
Breath Volume	2.515	21:51
Air Blank	0.000	21:51
Control Test	0.079	21:52
Air Blank	0.000	21:52
Diagnostics Check	OK	21:52

Data Download Date/Time: 08/03/2018 14:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAEZ HERNANDEZ,DENYS,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: PAEZ HERNANDEZ,DENYS
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A1
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: BJORKLUND

DATE/TIME OF TEST: 07/22/2018 04:35:48

Diagnostics Check	OK	04:27	
Air Blank	0.000	04:27	
Control Test	0.080	04:27	
Air Blank	0.000	04:28	
Subject Sample	# 1	04:31	No Sample Provided
Breath Volume	0.000	04:31	
Air Blank	0.000	04:31	
Air Blank	0.000	04:33	
Subject Sample	# 2	04:34	Subject Test Refused
Breath Volume	0.000	04:34	
Air Blank	0.000	04:34	
Control Test	0.080	04:35	
Air Blank	0.000	04:35	
Diagnostics Check	OK	04:35	

Data Download Date/Time: 08/03/2018 14:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAEZ HERNANDEZ,DENYS,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: PAEZ HERNANDEZ,DENYS
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A1
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: RILEY

DATE/TIME OF TEST: 07/26/2018 04:58:54

Diagnostics Check	OK	04:50	
Air Blank	0.000	04:50	
Control Test	0.080	04:51	
Air Blank	0.000	04:51	
Subject Sample	# 1	04:54	No Sample Provided
Breath Volume	0.000	04:54	
Air Blank	0.000	04:55	
Air Blank	0.000	04:57	
Subject Sample	# 2	04:57	Subject Test Refused
Breath Volume	0.000	04:57	
Air Blank	0.000	04:57	
Control Test	0.080	04:58	
Air Blank	0.000	04:58	
Diagnostics Check	OK	04:58	

Data Download Date/Time: 08/03/2018 14:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURDOCK,TROY,J
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: MURDOCK,TROY
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: MIZRACHI

DATE/TIME OF TEST: 07/21/2018 02:17:08

Diagnostics Check	OK	02:10
Air Blank	0.000	02:10
Control Test	0.078	02:11
Air Blank	0.000	02:11
Subject Sample # 1	0.178	02:12
Breath Volume	1.375	02:12
Air Blank	0.000	02:12
Air Blank	0.000	02:14
Subject Sample # 2	0.186	02:15
Breath Volume	2.050	02:15
Air Blank	0.000	02:16
Control Test	0.079	02:16
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

Data Download Date/Time: 07/25/2018 17:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,MATTHEW,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CARTER,MATTHEW
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 07/02/2018 02:47:00

Diagnostics Check	OK	02:43	
Air Blank	0.000	02:44	
Control Test	0.079	02:44	
Air Blank	0.000	02:44	
Subject Sample	# 1 0.239	02:45	
Breath Volume	1.105	02:45	
Air Blank		02:45	RFI Detect
Air Blank	0.000	02:46	
Control Test	0.078	02:46	
Air Blank	0.000	02:46	
Diagnostics Check	OK	02:47	

Data Download Date/Time: 07/11/2018 10:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,MATTHEW,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CARTER,MATTHEW
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 07/02/2018 02:55:05

Diagnostics Check	OK	02:48
Air Blank	0.000	02:49
Control Test	0.079	02:49
Air Blank	0.000	02:49
Subject Sample # 1	0.238	02:50
Breath Volume	1.234	02:50
Air Blank	0.000	02:51
Air Blank	0.000	02:52
Subject Sample # 2	0.234	02:53
Breath Volume	1.410	02:53
Air Blank	0.000	02:54
Control Test	0.079	02:54
Air Blank	0.000	02:54
Diagnostics Check	OK	02:55

Data Download Date/Time: 07/11/2018 10:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,MATTHEW,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SAIA,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: SHANNON

DATE/TIME OF TEST: 07/02/2018 03:27:51

Diagnostics Check	OK	03:21
Air Blank	0.000	03:21
Control Test	0.078	03:22
Air Blank	0.000	03:22
Subject Sample # 1	0.145	03:23
Breath Volume	1.621	03:23
Air Blank	0.000	03:23
Air Blank	0.000	03:25
Subject Sample # 2	0.144	03:26
Breath Volume	1.308	03:26
Air Blank	0.000	03:26
Control Test	0.079	03:27
Air Blank	0.000	03:27
Diagnostics Check	OK	03:27

Data Download Date/Time: 07/11/2018 10:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE, RONALD, W
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCKENZIE, RONALD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 15:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: BADGETT

DATE/TIME OF TEST: 07/03/2018 16:27:28

Diagnostics Check	OK	16:21
Air Blank	0.000	16:21
Control Test	0.080	16:22
Air Blank	0.000	16:22
Subject Sample # 1	0.008	16:23
Breath Volume	3.472	16:23
Air Blank	0.000	16:23
Air Blank	0.000	16:25
Subject Sample # 2	0.000	16:26
Breath Volume	2.683	16:26
Air Blank	0.000	16:26
Control Test	0.078	16:26
Air Blank	0.000	16:27
Diagnostics Check	OK	16:27

Data Download Date/Time: 07/11/2018 10:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAIA,MICHAEL,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ACRI,MICHEAL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: MURIN

DATE/TIME OF TEST: 07/05/2018 05:36:51

Diagnostics Check	OK	05:28
Air Blank	0.000	05:28
Control Test	0.078	05:28
Air Blank	0.000	05:29
Subject Sample # 1	0.236	05:32
Breath Volume	3.351	05:32
Air Blank	0.000	05:33
Air Blank	0.000	05:34
Subject Sample # 2	0.243	05:35
Breath Volume	3.437	05:35
Air Blank	0.000	05:35
Control Test	0.078	05:36
Air Blank	0.000	05:36
Diagnostics Check	OK	05:36

Data Download Date/Time: 07/11/2018 10:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUM,ZACKARY,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WALDO,AARON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: KORICA

DATE/TIME OF TEST: 07/06/2018 00:26:47

Diagnostics Check	OK	00:20
Air Blank	0.000	00:20
Control Test	0.078	00:20
Air Blank	0.000	00:21
Subject Sample # 1	0.178	00:22
Breath Volume	1.578	00:22
Air Blank	0.000	00:22
Air Blank	0.000	00:24
Subject Sample # 2	0.177	00:25
Breath Volume	3.578	00:25
Air Blank	0.000	00:25
Control Test	0.079	00:26
Air Blank	0.000	00:26
Diagnostics Check	OK	00:26

Data Download Date/Time: 07/11/2018 10:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAIA,MICHAEL,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SAIA,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: REILMANN

DATE/TIME OF TEST: 07/09/2018 05:01:33

Diagnostics Check	OK	04:55
Air Blank	0.000	04:55
Control Test	0.079	04:55
Air Blank	0.000	04:56
Subject Sample # 1	0.158	04:57
Breath Volume	2.308	04:57
Air Blank	0.000	04:57
Air Blank	0.000	04:59
Subject Sample # 2	0.159	05:00
Breath Volume	2.218	05:00
Air Blank	0.000	05:00
Control Test	0.078	05:00
Air Blank	0.000	05:01
Diagnostics Check	OK	05:01

Data Download Date/Time: 07/11/2018 10:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,M,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WALDO,AARON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: MCMURDOCK

DATE/TIME OF TEST: 07/11/2018 21:39:22

Diagnostics Check	OK	21:32
Air Blank	0.000	21:33
Control Test	0.078	21:33
Air Blank	0.000	21:34
Subject Sample # 1	0.280	21:35
Breath Volume	2.550	21:35
Air Blank	0.000	21:35
Air Blank	0.000	21:37
Subject Sample # 2	0.269	21:37
Breath Volume	2.078	21:37
Air Blank	0.000	21:38
Control Test	0.076	21:38
Air Blank	0.000	21:39
Diagnostics Check	OK	21:39

Data Download Date/Time: 08/21/2018 09:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDO,AARON,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KARAYIANES,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: CHILEL LOPEZ

DATE/TIME OF TEST: 07/12/2018 02:53:41

Diagnostics Check	OK	02:45
Air Blank	0.000	02:46
Control Test	0.077	02:46
Air Blank	0.000	02:47
Subject Sample # 1	0.230	02:48
Breath Volume	3.062	02:48
Air Blank	0.000	02:49
Air Blank	0.000	02:51
Subject Sample # 2	0.228	02:52
Breath Volume	2.437	02:52
Air Blank	0.000	02:52
Control Test	0.078	02:53
Air Blank	0.000	02:53
Diagnostics Check	OK	02:53

Data Download Date/Time: 08/21/2018 09:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUM,ZACKARY,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WALDO,AARON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: MIDDLETON

DATE/TIME OF TEST: 07/13/2018 02:31:03

Diagnostics Check	OK	02:22
Air Blank	0.000	02:23
Control Test	0.078	02:23
Air Blank	0.000	02:23
Subject Sample # 1	0.154	02:24
Breath Volume	1.460	02:24
Air Blank	0.000	02:25
Air Blank	0.000	02:27
Subject Sample # 2	0.149	02:29
Breath Volume	1.308	02:29
Air Blank	0.000	02:30
Control Test	0.079	02:30
Air Blank	0.000	02:30
Diagnostics Check	OK	02:31

Data Download Date/Time: 08/21/2018 09:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRONZ,BRIAN,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WALDO,AARON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: SCHULTE

DATE/TIME OF TEST: 07/13/2018 23:58:31

Diagnostics Check	OK	23:54	
Air Blank	0.000	23:54	
Control Test	0.078	23:55	
Air Blank	0.000	23:55	
Subject Sample	# 1	23:57	Subject Test Refused
Breath Volume	0.000	23:57	
Air Blank	0.000	23:57	
Control Test	0.079	23:57	
Air Blank	0.000	23:58	
Diagnostics Check	OK	23:58	

Data Download Date/Time: 08/21/2018 09:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAQUERO,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HAUSAM,MATTHEW
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: BAQUERO

DATE/TIME OF TEST: 07/14/2018 03:46:39

Diagnostics Check	OK	03:42	
Air Blank	0.000	03:43	
Control Test	0.078	03:43	
Air Blank	0.000	03:44	
Subject Sample	# 1	03:45	Subject Test Refused
Breath Volume	0.000	03:45	
Air Blank	0.000	03:45	
Control Test	0.079	03:46	
Air Blank	0.000	03:46	
Diagnostics Check	OK	03:46	

Data Download Date/Time: 08/21/2018 09:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CRUM,ZACKARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/15/2018 00:41:29

Diagnostics Check	OK	00:38	
Air Blank	0.000	00:39	
Control Test	0.079	00:39	
Air Blank	0.000	00:39	
Subject Sample	# 1	00:40	Subject Test Refused
Breath Volume	0.000	00:40	
Air Blank	0.000	00:40	
Control Test	0.079	00:40	
Air Blank	0.000	00:41	
Diagnostics Check	OK	00:41	

Data Download Date/Time: 08/21/2018 09:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CRUM,ZACKARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: PETTAY

DATE/TIME OF TEST: 07/15/2018 03:12:14

Diagnostics Check	OK	03:05
Air Blank	0.000	03:05
Control Test	0.079	03:05
Air Blank	0.000	03:06
Subject Sample # 1	0.091	03:07
Breath Volume	2.007	03:07
Air Blank	0.000	03:08
Air Blank	0.000	03:09
Subject Sample # 2	0.087	03:10
Breath Volume	1.703	03:10
Air Blank	0.000	03:11
Control Test	0.078	03:11
Air Blank	0.000	03:12
Diagnostics Check	OK	03:12

Data Download Date/Time: 08/21/2018 09:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,EMILY,K
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,KRISTEN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: TALLON

DATE/TIME OF TEST: 07/16/2018 22:24:40

Diagnostics Check	OK	22:21	
Air Blank	0.000	22:22	
Control Test	0.078	22:22	
Air Blank	0.000	22:22	
Subject Sample	# 1	22:23	Subject Test Refused
Breath Volume	0.000	22:23	
Air Blank	0.000	22:23	
Control Test	0.079	22:24	
Air Blank	0.000	22:24	
Diagnostics Check	OK	22:24	

Data Download Date/Time: 08/21/2018 09:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUM,ZACKARY,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA,V
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 07/18/2018 21:01:47

Diagnostics Check	OK	20:45	
Air Blank	0.000	20:46	
Control Test	0.080	20:46	
Air Blank	0.000	20:47	
Subject Sample	# 1 0.000	20:50	Volume Not Met
Breath Volume	0.367	20:50	
Air Blank	0.000	20:50	
Air Blank	0.000	20:52	
Subject Sample	# 2 0.000	20:54	
Breath Volume	1.578	20:54	
Air Blank	0.000	20:55	
Air Blank	0.000	20:57	
Subject Sample	# 3 0.000	21:00	Volume Not Met
Breath Volume	0.261	21:00	
Air Blank	0.000	21:00	
Control Test	0.080	21:01	
Air Blank	0.000	21:01	
Diagnostics Check	OK	21:01	

Data Download Date/Time: 08/21/2018 09:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUM,ZACKARY,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA,V
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 07/18/2018 21:16:14

Diagnostics Check	OK	21:05
Air Blank	0.000	21:05
Control Test	0.079	21:05
Air Blank	0.000	21:06
Subject Sample # 1	0.000	21:09
Breath Volume	2.050	21:09
Air Blank	0.000	21:10
Air Blank	0.000	21:12
Subject Sample # 2	0.000	21:14
Breath Volume	2.550	21:14
Air Blank	0.000	21:15
Control Test	0.080	21:15
Air Blank	0.000	21:16
Diagnostics Check	OK	21:16

Data Download Date/Time: 08/21/2018 09:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KARAYIANES,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 07/20/2018 23:17:11

Diagnostics Check	OK	23:10
Air Blank	0.000	23:11
Control Test	0.079	23:11
Air Blank	0.000	23:12
Subject Sample # 1	0.274	23:12
Breath Volume	1.519	23:12
Air Blank	0.000	23:13
Air Blank	0.000	23:15
Subject Sample # 2	0.271	23:15
Breath Volume	2.156	23:15
Air Blank	0.000	23:16
Control Test	0.078	23:16
Air Blank	0.000	23:17
Diagnostics Check	OK	23:17

Data Download Date/Time: 08/21/2018 09:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,MATTHEW,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SAIA,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: MALLORY

DATE/TIME OF TEST: 07/23/2018 03:50:21

Diagnostics Check	OK	03:43
Air Blank	0.000	03:44
Control Test	0.079	03:44
Air Blank	0.000	03:45
Subject Sample # 1	0.145	03:45
Breath Volume	2.414	03:45
Air Blank	0.000	03:46
Air Blank	0.000	03:48
Subject Sample # 2	0.145	03:48
Breath Volume	2.171	03:48
Air Blank	0.000	03:49
Control Test	0.078	03:49
Air Blank	0.000	03:50
Diagnostics Check	OK	03:50

Data Download Date/Time: 08/21/2018 09:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDO,AARON,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KARAYIANES,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: HOLLOWAY

DATE/TIME OF TEST: 07/24/2018 22:18:46

Diagnostics Check	OK	22:12
Air Blank	0.000	22:13
Control Test	0.078	22:13
Air Blank	0.000	22:13
Subject Sample # 1	0.110	22:14
Breath Volume	2.636	22:14
Air Blank	0.000	22:15
Air Blank	0.000	22:16
Subject Sample # 2	0.108	22:17
Breath Volume	2.437	22:17
Air Blank	0.000	22:17
Control Test	0.078	22:18
Air Blank	0.000	22:18
Diagnostics Check	OK	22:18

Data Download Date/Time: 08/21/2018 09:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WALDO,AARON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: CALKINS JR

DATE/TIME OF TEST: 07/25/2018 00:00:38

Diagnostics Check	OK	23:50
Air Blank	0.000	23:50
Control Test	0.077	23:51
Air Blank	0.000	23:51
Subject Sample # 1	0.217	23:52
Breath Volume	3.457	23:52
Air Blank	0.000	23:53
Air Blank	0.000	23:55
Subject Sample # 2	0.196	23:55
Breath Volume	1.253	23:55
Air Blank	0.000	23:56
Air Blank	0.000	23:58
Subject Sample # 3	0.211	23:59
Breath Volume	2.839	23:59
Air Blank	0.000	23:59
Control Test	0.075	00:00
Air Blank	0.000	00:00
Diagnostics Check	OK	00:00

Data Download Date/Time: 08/21/2018 09:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA,VLADIMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: GUTIERREZ

DATE/TIME OF TEST: 07/25/2018 00:54:59

Diagnostics Check	OK	00:47
Air Blank	0.000	00:48
Control Test	0.079	00:48
Air Blank	0.000	00:49
Subject Sample # 1	0.190	00:49
Breath Volume	1.832	00:49
Air Blank	0.000	00:50
Air Blank	0.000	00:52
Subject Sample # 2	0.199	00:53
Breath Volume	1.839	00:53
Air Blank	0.000	00:54
Control Test	0.076	00:54
Air Blank	0.000	00:54
Diagnostics Check	OK	00:54

Data Download Date/Time: 08/21/2018 09:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KARAYIANES,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 07/28/2018 03:45:11

Diagnostics Check	OK	03:36	
Air Blank	0.000	03:36	
Control Test	0.078	03:36	
Air Blank	0.000	03:37	
Subject Sample	# 1	03:40	No Sample Provided
Breath Volume	0.000	03:40	
Air Blank	0.000	03:41	
Air Blank	0.000	03:43	
Subject Sample	# 2	03:43	Subject Test Refused
Breath Volume	0.000	03:43	
Air Blank	0.000	03:44	
Control Test	0.078	03:44	
Air Blank	0.000	03:45	
Diagnostics Check	OK	03:45	

Data Download Date/Time: 08/21/2018 09:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUM,ZACKARY,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KARAYIANES,M
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: BARNETT

DATE/TIME OF TEST: 07/28/2018 21:49:28

Diagnostics Check	OK	21:46	
Air Blank	0.000	21:46	
Control Test	0.078	21:47	
Air Blank	0.000	21:47	
Subject Sample	# 1	21:48	Interferent Detect
Breath Volume	3.640	21:48	
Air Blank	0.000	21:48	
Control Test	0.078	21:48	
Air Blank	0.000	21:49	
Diagnostics Check	OK	21:49	

Data Download Date/Time: 08/21/2018 09:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUM,ZACKARY,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KARAYIANES,M
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: BARNETT

DATE/TIME OF TEST: 07/28/2018 22:13:58

Diagnostics Check	OK	22:10	
Air Blank	0.000	22:11	
Control Test	0.078	22:11	
Air Blank	0.000	22:12	
Subject Sample	# 1	22:12	Interferent Detect
Breath Volume	1.824	22:12	
Air Blank	0.000	22:13	
Control Test	0.078	22:13	
Air Blank	0.000	22:13	
Diagnostics Check	OK	22:13	

Data Download Date/Time: 08/21/2018 09:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,SHANNON,R
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: MORRIS,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:11

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 07/01/2018 17:44:41

Diagnostics Check	OK	17:38
Air Blank	0.000	17:38
Control Test	0.078	17:38
Air Blank	0.000	17:39
Subject Sample # 1	0.095	17:40
Breath Volume	3.011	17:40
Air Blank	0.000	17:40
Air Blank	0.000	17:42
Subject Sample # 2	0.090	17:43
Breath Volume	1.539	17:43
Air Blank	0.000	17:43
Control Test	0.076	17:44
Air Blank	0.000	17:44
Diagnostics Check	OK	17:44

Data Download Date/Time: 08/03/2018 08:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIGGINBOTHAM,B,K
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: REYNOLDS,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:37

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 07/01/2018 20:02:25

Diagnostics Check	OK	19:59	
Air Blank	0.000	19:59	
Control Test	0.079	20:00	
Air Blank	0.000	20:00	
Subject Sample	# 1	20:01	Subject Test Refused
Breath Volume	0.000	20:01	
Air Blank	0.000	20:01	
Control Test	0.080	20:01	
Air Blank	0.000	20:02	
Diagnostics Check	OK	20:02	

Data Download Date/Time: 08/03/2018 08:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HEALY,GREGORY,B

BREATH TEST OPERATOR AGENCY: FHP Troop G

ARRESTING OFFICER: HEALY,GREGORY

ARRESTING OFFICER AGENCY: FHP Troop G

LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: Yes

DRY GAS STANDARD LOT #: 22817080A5

DRY GAS EXPIRY DATE: 10/05/2019

VIOLATION CODE: DUI

ARREST TIME: 21:40

GENDER: Female

OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: MULLIS-BLY

DATE/TIME OF TEST: 07/04/2018 23:34:42

Diagnostics Check	OK	23:28
Air Blank	0.000	23:29
Control Test	0.079	23:29
Air Blank	0.000	23:30
Subject Sample # 1	0.122	23:30
Breath Volume	1.742	23:30
Air Blank	0.000	23:30
Air Blank	0.000	23:32
Subject Sample # 2	0.122	23:33
Breath Volume	1.785	23:33
Air Blank	0.000	23:33
Control Test	0.080	23:34
Air Blank	0.000	23:34
Diagnostics Check	OK	23:34

Data Download Date/Time: 08/03/2018 08:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,D,R
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: GORDON,G
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: WILKERSON

DATE/TIME OF TEST: 07/05/2018 00:02:20

Diagnostics Check	OK	23:52
Air Blank	0.000	23:52
Control Test	0.079	23:53
Air Blank	0.000	23:53
Subject Sample # 2	0.116	23:55
Breath Volume	1.839	23:55
Air Blank	0.000	23:56
Air Blank	0.000	23:58
Subject Sample # 1	0.126	00:00
Breath Volume	2.808	00:00
Air Blank	0.000	00:01
Control Test	0.079	00:01
Air Blank	0.000	00:02
Diagnostics Check	OK	00:02

Data Download Date/Time: 08/03/2018 08:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARDONA,K,C
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: ENNIS,J
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 07/07/2018 03:28:01

Diagnostics Check	OK	03:25	
Air Blank	0.000	03:25	
Control Test	0.081	03:26	
Air Blank	0.000	03:26	
Subject Sample	# 1	03:26	Subject Test Refused
Breath Volume	0.000	03:26	
Air Blank	0.000	03:27	
Control Test	0.080	03:27	
Air Blank	0.000	03:27	
Diagnostics Check	OK	03:28	

Data Download Date/Time: 08/03/2018 08:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRES,J,L
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: CRIBBS

DATE/TIME OF TEST: 07/07/2018 03:39:16

Diagnostics Check	OK	03:33
Air Blank	0.000	03:33
Control Test	0.079	03:33
Air Blank	0.000	03:34
Subject Sample # 1	0.101	03:34
Breath Volume	1.882	03:34
Air Blank	0.000	03:35
Air Blank	0.000	03:37
Subject Sample # 2	0.097	03:37
Breath Volume	1.660	03:37
Air Blank	0.000	03:38
Control Test	0.079	03:38
Air Blank	0.000	03:39
Diagnostics Check	OK	03:39

Data Download Date/Time: 08/03/2018 08:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARDONA,K,C
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: DELACRUZ,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: FOSTER

DATE/TIME OF TEST: 07/07/2018 05:07:23

Diagnostics Check	OK	04:59	
Air Blank	0.000	04:59	
Control Test	0.079	04:59	
Air Blank	0.000	05:00	
Subject Sample	# 1	05:03	No Sample Provided
Breath Volume	0.000	05:03	
Air Blank	0.000	05:03	
Air Blank	0.000	05:05	
Subject Sample	# 2	05:06	Subject Test Refused
Breath Volume	0.000	05:06	
Air Blank	0.000	05:06	
Control Test	0.079	05:06	
Air Blank	0.000	05:07	
Diagnostics Check	OK	05:07	

Data Download Date/Time: 08/03/2018 08:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRES,J,L
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: KNIGHT,G
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: HUMBLE

DATE/TIME OF TEST: 07/09/2018 03:26:34

Diagnostics Check	OK	03:20
Air Blank	0.000	03:20
Control Test	0.081	03:21
Air Blank	0.000	03:21
Subject Sample # 1	0.187	03:22
Breath Volume	2.640	03:22
Air Blank	0.000	03:22
Air Blank	0.000	03:24
Subject Sample # 2	0.201	03:25
Breath Volume	2.695	03:25
Air Blank	0.000	03:25
Control Test	0.081	03:25
Air Blank	0.000	03:26
Diagnostics Check	OK	03:26

Data Download Date/Time: 08/03/2018 08:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCracken,J,D
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: DELACRUZ,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: HARDY

DATE/TIME OF TEST: 07/11/2018 22:48:05

Diagnostics Check	OK	22:41
Air Blank	0.000	22:41
Control Test	0.079	22:42
Air Blank	0.000	22:42
Subject Sample # 1	0.170	22:43
Breath Volume	1.554	22:43
Air Blank	0.000	22:43
Air Blank	0.000	22:45
Subject Sample # 2	0.173	22:46
Breath Volume	1.351	22:46
Air Blank	0.000	22:47
Control Test	0.079	22:47
Air Blank	0.000	22:48
Diagnostics Check	OK	22:48

Data Download Date/Time: 08/03/2018 08:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRES,J,L
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: SHOOK,J
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: STEVENSON

DATE/TIME OF TEST: 07/11/2018 23:00:43

Diagnostics Check	OK	22:54
Air Blank	0.000	22:54
Control Test	0.080	22:55
Air Blank	0.000	22:55
Subject Sample # 1	0.163	22:56
Breath Volume	2.476	22:56
Air Blank	0.000	22:56
Air Blank	0.000	22:58
Subject Sample # 2	0.161	22:59
Breath Volume	1.671	22:59
Air Blank	0.000	22:59
Control Test	0.079	23:00
Air Blank	0.000	23:00
Diagnostics Check	OK	23:00

Data Download Date/Time: 08/03/2018 08:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARDONA,K,C
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: HULL

DATE/TIME OF TEST: 07/12/2018 01:28:52

Diagnostics Check	OK	01:26	
Air Blank	0.000	01:26	
Control Test	0.079	01:26	
Air Blank	0.000	01:27	
Subject Sample	# 1	01:27	Subject Test Refused
Breath Volume	0.000	01:27	
Air Blank	0.000	01:27	
Control Test	0.079	01:28	
Air Blank	0.000	01:28	
Diagnostics Check	OK	01:28	

Data Download Date/Time: 08/03/2018 08:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRES,J,L
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: MCKABA

DATE/TIME OF TEST: 07/12/2018 01:50:52

Diagnostics Check	OK	01:44
Air Blank	0.000	01:44
Control Test	0.080	01:45
Air Blank	0.000	01:45
Subject Sample # 1	0.000	01:46
Breath Volume	2.070	01:46
Air Blank	0.000	01:46
Air Blank	0.000	01:48
Subject Sample # 2	0.000	01:49
Breath Volume	2.261	01:49
Air Blank	0.000	01:49
Control Test	0.079	01:50
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

Data Download Date/Time: 08/03/2018 08:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARDONA,K,C
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 07/13/2018 03:41:12

Diagnostics Check	OK	03:34
Air Blank	0.000	03:35
Control Test	0.078	03:35
Air Blank	0.000	03:35
Subject Sample # 1	0.091	03:36
Breath Volume	1.738	03:36
Air Blank	0.000	03:37
Air Blank	0.000	03:39
Subject Sample # 2	0.088	03:39
Breath Volume	1.402	03:39
Air Blank	0.000	03:40
Control Test	0.079	03:40
Air Blank	0.000	03:41
Diagnostics Check	OK	03:41

Data Download Date/Time: 08/03/2018 08:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEDDERLY,GARRETT,A
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: DELACRUZ,C
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:09

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 07/17/2018 06:46:57

Diagnostics Check	OK	06:44	
Air Blank	0.000	06:44	
Control Test	0.079	06:44	
Air Blank	0.000	06:45	
Subject Sample	# 1	06:45	Subject Test Refused
Breath Volume	0.000	06:45	
Air Blank	0.000	06:46	
Control Test	0.079	06:46	
Air Blank	0.000	06:46	
Diagnostics Check	OK	06:46	

Data Download Date/Time: 08/03/2018 08:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRES,J,L
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: MCCAIN

DATE/TIME OF TEST: 07/17/2018 23:13:06

Diagnostics Check	OK	23:10	
Air Blank	0.000	23:10	
Control Test	0.080	23:11	
Air Blank	0.000	23:11	
Subject Sample	# 1	23:11	Subject Test Refused
Breath Volume	0.000	23:11	
Air Blank	0.000	23:12	
Control Test	0.080	23:12	
Air Blank	0.000	23:13	
Diagnostics Check	OK	23:13	

Data Download Date/Time: 08/03/2018 08:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,S,R
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: SWEAT,TIM
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 13:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:01

SUBJECT'S LAST NAME: PEAKE

DATE/TIME OF TEST: 07/19/2018 15:49:11

Diagnostics Check	OK	15:39
Air Blank	0.000	15:40
Control Test	0.080	15:40
Air Blank	0.000	15:41
Subject Sample # 1	0.292	15:41
Breath Volume	2.796	15:41
Air Blank	0.000	15:42
Air Blank	0.000	15:44
Subject Sample # 2	0.267	15:44
Breath Volume	1.570	15:44
Air Blank	0.000	15:45
Air Blank	0.000	15:47
Subject Sample # 3	0.249	15:47
Breath Volume	1.230	15:47
Air Blank	0.000	15:48
Control Test	0.076	15:48
Air Blank	0.000	15:49
Diagnostics Check	OK	15:49

Data Download Date/Time: 08/03/2018 08:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBB,CATHERINE,D
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: DUGGAN,L
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 07/20/2018 00:39:25

Diagnostics Check	OK	00:28	
Air Blank	0.000	00:28	
Control Test	0.080	00:28	
Air Blank	0.000	00:29	
Subject Sample	# 1	00:32	No Sample Provided
Breath Volume	0.000	00:32	
Air Blank	0.000	00:32	
Air Blank	0.000	00:34	
Subject Sample	# 2	00:38	No Sample Provided
Breath Volume	0.000	00:38	
Air Blank	0.000	00:38	
Control Test	0.079	00:38	
Air Blank	0.000	00:39	
Diagnostics Check	OK	00:39	

Data Download Date/Time: 08/03/2018 08:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCracken,J,D
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: FROST

DATE/TIME OF TEST: 07/20/2018 23:08:29

Diagnostics Check	OK	23:02
Air Blank	0.000	23:02
Control Test	0.080	23:03
Air Blank	0.000	23:03
Subject Sample # 1	0.151	23:04
Breath Volume	1.117	23:04
Air Blank	0.000	23:04
Air Blank	0.000	23:06
Subject Sample # 2	0.143	23:06
Breath Volume	1.753	23:06
Air Blank	0.000	23:07
Control Test	0.078	23:07
Air Blank	0.000	23:08
Diagnostics Check	OK	23:08

Data Download Date/Time: 08/03/2018 08:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARDONA,K,C
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: KNIGHT,G
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 07/21/2018 23:30:18

Diagnostics Check	OK	23:23
Air Blank	0.000	23:23
Control Test	0.080	23:24
Air Blank	0.000	23:24
Subject Sample # 1	0.116	23:25
Breath Volume	1.906	23:25
Air Blank	0.000	23:26
Air Blank	0.000	23:28
Subject Sample # 2	0.111	23:28
Breath Volume	1.445	23:28
Air Blank	0.000	23:29
Control Test	0.079	23:29
Air Blank	0.000	23:30
Diagnostics Check	OK	23:30

Data Download Date/Time: 08/03/2018 08:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEALY,GREGORY,B
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HEALY,GREGORY
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: CLEGG

DATE/TIME OF TEST: 07/22/2018 02:56:24

Diagnostics Check	OK	02:50
Air Blank	0.000	02:50
Control Test	0.079	02:50
Air Blank	0.000	02:51
Subject Sample # 1	0.206	02:52
Breath Volume	4.085	02:52
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample # 2	0.206	02:54
Breath Volume	1.351	02:54
Air Blank	0.000	02:55
Control Test	0.080	02:55
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 08/03/2018 08:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,S,R
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HILLIARD,N
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 15:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:43

SUBJECT'S LAST NAME: LIVINGSTONE

DATE/TIME OF TEST: 07/24/2018 17:19:34

Diagnostics Check	OK	17:13
Air Blank	0.000	17:13
Control Test	0.078	17:13
Air Blank	0.000	17:14
Subject Sample # 1	0.236	17:15
Breath Volume	3.375	17:15
Air Blank	0.000	17:15
Air Blank	0.000	17:17
Subject Sample # 2	0.234	17:18
Breath Volume	3.445	17:18
Air Blank	0.000	17:18
Control Test	0.079	17:19
Air Blank	0.000	17:19
Diagnostics Check	OK	17:19

Data Download Date/Time: 08/03/2018 08:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIGGINBOTHAM,B,K
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: BALLARD,J
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:01

SUBJECT'S LAST NAME: MERCER

DATE/TIME OF TEST: 07/25/2018 05:29:12

Diagnostics Check	OK	05:22
Air Blank	0.000	05:23
Control Test	0.080	05:23
Air Blank	0.000	05:23
Subject Sample # 1	0.000	05:24
Breath Volume	2.593	05:24
Air Blank	0.000	05:25
Air Blank	0.000	05:27
Subject Sample # 2	0.000	05:27
Breath Volume	1.128	05:27
Air Blank	0.000	05:28
Control Test	0.080	05:28
Air Blank	0.000	05:29
Diagnostics Check	OK	05:29

Data Download Date/Time: 08/03/2018 08:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARDONA,K,C
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: KNIGHT,G
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: JAMES

DATE/TIME OF TEST: 07/26/2018 04:50:45

Diagnostics Check	OK	04:47	
Air Blank	0.000	04:48	
Control Test	0.080	04:48	
Air Blank	0.000	04:49	
Subject Sample	# 1	04:49	Subject Test Refused
Breath Volume	0.000	04:49	
Air Blank	0.000	04:49	
Control Test	0.079	04:50	
Air Blank	0.000	04:50	
Diagnostics Check	OK	04:50	

Data Download Date/Time: 08/03/2018 08:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARDONA,K,C
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: KNIGHT,G
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: PRINGLE

DATE/TIME OF TEST: 07/26/2018 21:13:20

Diagnostics Check	OK	21:01
Air Blank	0.000	21:01
Control Test	0.079	21:01
Air Blank	0.000	21:02
Subject Sample # 1	0.150	21:03
Breath Volume	2.046	21:03
Air Blank	0.000	21:03
Air Blank	0.000	21:05
Subject Sample # 2	0.124	21:07
Breath Volume	1.894	21:07
Air Blank	0.000	21:08
Air Blank	0.000	21:10
Subject Sample # 3	0.133	21:11
Breath Volume	1.492	21:11
Air Blank	0.000	21:12
Control Test	0.079	21:12
Air Blank	0.000	21:13
Diagnostics Check	OK	21:13

Data Download Date/Time: 08/03/2018 08:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,S,R
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: PAOLILLO,K
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:25

SUBJECT'S LAST NAME: NABORS

DATE/TIME OF TEST: 07/27/2018 17:00:15

Diagnostics Check	OK	16:53
Air Blank	0.000	16:54
Control Test	0.079	16:54
Air Blank	0.000	16:55
Subject Sample # 1	0.307	16:55
Breath Volume	2.148	16:55
Air Blank	0.000	16:56
Air Blank	0.000	16:58
Subject Sample # 2	0.317	16:58
Breath Volume	2.281	16:58
Air Blank	0.000	16:59
Control Test	0.079	16:59
Air Blank	0.000	17:00
Diagnostics Check	OK	17:00

Data Download Date/Time: 08/03/2018 08:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATHIS,AUSTIN,J
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: DUGGAN,L
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: ROWLAND

DATE/TIME OF TEST: 07/28/2018 01:40:06

Diagnostics Check	OK	01:31
Air Blank	0.000	01:32
Control Test	0.080	01:32
Air Blank	0.000	01:33
Subject Sample # 1	0.000	01:34
Breath Volume	1.128	01:34
Air Blank	0.000	01:34
Air Blank	0.000	01:36
Subject Sample # 2	0.000	01:38
Breath Volume	1.117	01:38
Air Blank	0.000	01:39
Control Test	0.079	01:39
Air Blank	0.000	01:40
Diagnostics Check	OK	01:40

Data Download Date/Time: 08/03/2018 08:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,D,R
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: ASARO,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:23

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/28/2018 19:51:01

Diagnostics Check	OK	19:44
Air Blank	0.000	19:44
Control Test	0.080	19:44
Air Blank	0.000	19:45
Subject Sample # 1	0.221	19:46
Breath Volume	3.714	19:46
Air Blank	0.000	19:46
Air Blank	0.000	19:48
Subject Sample # 2	0.216	19:49
Breath Volume	3.851	19:49
Air Blank	0.000	19:50
Control Test	0.079	19:50
Air Blank	0.000	19:50
Diagnostics Check	OK	19:51

Data Download Date/Time: 08/03/2018 08:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEDDERLY,G,A
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: ENNIS,J
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:18

SUBJECT'S LAST NAME: MOZO

DATE/TIME OF TEST: 07/31/2018 17:43:59

Diagnostics Check	OK	17:40	
Air Blank	0.000	17:41	
Control Test	0.079	17:41	
Air Blank	0.000	17:41	
Subject Sample	# 1	17:42	Subject Test Refused
Breath Volume	0.000	17:42	
Air Blank	0.000	17:43	
Control Test	0.080	17:43	
Air Blank	0.000	17:43	
Diagnostics Check	OK	17:43	

Data Download Date/Time: 08/03/2018 08:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARLOW,ROYCE,B
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SMITH,RYAN
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: BUI
ARREST TIME: 19:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:28

SUBJECT'S LAST NAME: PICKERING

DATE/TIME OF TEST: 07/01/2018 21:58:29

Diagnostics Check	OK	21:52
Air Blank	0.000	21:52
Control Test	0.079	21:53
Air Blank	0.000	21:53
Subject Sample # 1	0.011	21:54
Breath Volume	1.437	21:54
Air Blank	0.000	21:54
Air Blank	0.000	21:56
Subject Sample # 2	0.011	21:57
Breath Volume	1.414	21:57
Air Blank	0.000	21:57
Control Test	0.079	21:57
Air Blank	0.000	21:58
Diagnostics Check	OK	21:58

Data Download Date/Time: 07/28/2018 22:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GLASCO,GREGORY,S
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: COOPER,GARY
ARRESTING OFFICER AGENCY: Lake Placid PD
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: KRICK

DATE/TIME OF TEST: 07/10/2018 22:56:34

Diagnostics Check	OK	22:50
Air Blank	0.000	22:50
Control Test	0.078	22:50
Air Blank	0.000	22:51
Subject Sample # 1	0.175	22:51
Breath Volume	2.117	22:51
Air Blank	0.000	22:52
Air Blank	0.000	22:54
Subject Sample # 2	0.166	22:54
Breath Volume	1.312	22:54
Air Blank	0.000	22:55
Control Test	0.078	22:55
Air Blank	0.000	22:56
Diagnostics Check	OK	22:56

Data Download Date/Time: 07/28/2018 22:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WORTH,KEVIN,J
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: RHOADEN,D
ARRESTING OFFICER AGENCY: Lake Placid PD
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: ALMEIDA

DATE/TIME OF TEST: 07/13/2018 03:39:21

Diagnostics Check	OK	03:32
Air Blank	0.000	03:33
Control Test	0.079	03:33
Air Blank	0.000	03:34
Subject Sample # 1	0.207	03:34
Breath Volume	2.687	03:34
Air Blank	0.000	03:35
Air Blank	0.000	03:37
Subject Sample # 2	0.201	03:37
Breath Volume	2.886	03:37
Air Blank	0.000	03:38
Control Test	0.079	03:38
Air Blank	0.000	03:39
Diagnostics Check	OK	03:39

Data Download Date/Time: 07/28/2018 22:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,C,L
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SALVADOR,E
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: LAVENDER

DATE/TIME OF TEST: 07/14/2018 03:24:48

Diagnostics Check	OK	03:18
Air Blank	0.000	03:19
Control Test	0.078	03:19
Air Blank	0.000	03:19
Subject Sample # 1	0.112	03:20
Breath Volume	2.644	03:20
Air Blank	0.000	03:21
Air Blank	0.000	03:22
Subject Sample # 2	0.108	03:23
Breath Volume	1.953	03:23
Air Blank	0.000	03:23
Control Test	0.079	03:24
Air Blank	0.000	03:24
Diagnostics Check	OK	03:24

Data Download Date/Time: 07/28/2018 22:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AYALA,JOSE,L
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SALVADOR,ELDER
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: KASSAM

DATE/TIME OF TEST: 07/17/2018 23:15:15

Diagnostics Check	OK	23:11	
Air Blank	0.000	23:12	
Control Test	0.079	23:12	
Air Blank	0.000	23:12	
Subject Sample	# 1	23:13	Subject Test Refused
Breath Volume	0.000	23:13	
Air Blank	0.000	23:14	
Control Test	0.078	23:14	
Air Blank	0.000	23:15	
Diagnostics Check	OK	23:15	

Data Download Date/Time: 07/28/2018 22:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WORTH,K,J
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: AMAYA,P
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: HOOVER

DATE/TIME OF TEST: 07/20/2018 23:21:36

Diagnostics Check	OK	23:15
Air Blank	0.000	23:15
Control Test	0.079	23:16
Air Blank	0.000	23:16
Subject Sample # 1	0.185	23:17
Breath Volume	1.539	23:17
Air Blank	0.000	23:17
Air Blank	0.000	23:19
Subject Sample # 2	0.196	23:20
Breath Volume	1.816	23:20
Air Blank	0.000	23:20
Control Test	0.078	23:21
Air Blank	0.000	23:21
Diagnostics Check	OK	23:21

Data Download Date/Time: 07/28/2018 22:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WORTH,K,J
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: BECERRA,J
ARRESTING OFFICER AGENCY: Lake Placid PD
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 07/21/2018 03:27:39

Diagnostics Check	OK	03:20
Air Blank	0.000	03:20
Control Test	0.078	03:21
Air Blank	0.000	03:21
Subject Sample # 1	0.057	03:22
Breath Volume	2.019	03:22
Air Blank	0.000	03:22
Air Blank	0.000	03:24
Subject Sample # 2	0.057	03:26
Breath Volume	1.753	03:26
Air Blank	0.000	03:26
Control Test	0.079	03:27
Air Blank	0.000	03:27
Diagnostics Check	OK	03:27

Data Download Date/Time: 07/28/2018 22:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,REBECCA,C
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: MARIN,A
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 10:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:26

SUBJECT'S LAST NAME: ESPARZA

DATE/TIME OF TEST: 07/21/2018 11:59:26

Diagnostics Check	OK	11:53
Air Blank	0.000	11:53
Control Test	0.078	11:54
Air Blank	0.000	11:54
Subject Sample # 1	0.000	11:55
Breath Volume	2.343	11:55
Air Blank	0.000	11:55
Air Blank	0.000	11:57
Subject Sample # 2	0.000	11:58
Breath Volume	1.910	11:58
Air Blank	0.000	11:58
Control Test	0.079	11:58
Air Blank	0.000	11:59
Diagnostics Check	OK	11:59

Data Download Date/Time: 07/28/2018 22:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,SUSAN,L
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: BOGUSKY,D
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: WELCH

DATE/TIME OF TEST: 07/25/2018 01:51:18

Diagnostics Check	OK	01:48	
Air Blank	0.000	01:48	
Control Test	0.079	01:49	
Air Blank	0.000	01:49	
Subject Sample	# 1	01:49	Subject Test Refused
Breath Volume	0.000	01:49	
Air Blank	0.000	01:50	
Control Test	0.078	01:50	
Air Blank	0.000	01:51	
Diagnostics Check	OK	01:51	

Data Download Date/Time: 07/28/2018 22:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PECK,JOHN,R
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: MARIN,ADRAIN
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 07:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:44

SUBJECT'S LAST NAME: MCCRAY

DATE/TIME OF TEST: 07/26/2018 09:11:24

Diagnostics Check	OK	09:05
Air Blank	0.000	09:05
Control Test	0.079	09:05
Air Blank	0.000	09:06
Subject Sample # 1	0.000	09:07
Breath Volume	2.523	09:07
Air Blank	0.000	09:07
Air Blank	0.000	09:09
Subject Sample # 2	0.000	09:10
Breath Volume	2.406	09:10
Air Blank	0.000	09:10
Control Test	0.078	09:10
Air Blank	0.000	09:11
Diagnostics Check	OK	09:11

Data Download Date/Time: 07/28/2018 22:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAUGHN,DALE,E
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SALVADOR,ELDER
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: CARRANZA PEREZ

DATE/TIME OF TEST: 07/27/2018 22:20:44

Diagnostics Check	OK	22:14
Air Blank	0.000	22:14
Control Test	0.079	22:14
Air Blank	0.000	22:15
Subject Sample # 1	0.131	22:16
Breath Volume	2.445	22:16
Air Blank	0.000	22:16
Air Blank	0.000	22:18
Subject Sample # 2	0.129	22:19
Breath Volume	2.636	22:19
Air Blank	0.000	22:19
Control Test	0.079	22:20
Air Blank	0.000	22:20
Diagnostics Check	OK	22:20

Data Download Date/Time: 07/28/2018 22:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAUGHN,DALE,E
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SALVADOR,ELDER
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: HARKNESS

DATE/TIME OF TEST: 07/28/2018 02:42:50

Diagnostics Check	OK	02:35
Air Blank	0.000	02:36
Control Test	0.079	02:36
Air Blank	0.000	02:36
Subject Sample # 1	0.229	02:38
Breath Volume	1.425	02:38
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample # 2	0.222	02:41
Breath Volume	1.601	02:41
Air Blank	0.000	02:41
Control Test	0.079	02:42
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 07/28/2018 22:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAUGHN, DALE, E
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: COOPER, OFFICER
ARRESTING OFFICER AGENCY: Lake Placid PD
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: KELLEY

DATE/TIME OF TEST: 07/28/2018 04:01:35

Diagnostics Check	OK	03:58	
Air Blank	0.000	03:59	
Control Test	0.079	03:59	
Air Blank	0.000	03:59	
Subject Sample	# 1	04:00	Subject Test Refused
Breath Volume	0.000	04:00	
Air Blank	0.000	04:00	
Control Test	0.079	04:00	
Air Blank	0.000	04:01	
Diagnostics Check	OK	04:01	

Data Download Date/Time: 07/28/2018 22:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNELL,C,J
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: FRANZE,F
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 14:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:20

SUBJECT'S LAST NAME: ZAPATA

DATE/TIME OF TEST: 07/28/2018 15:44:48

Diagnostics Check	OK	15:41	
Air Blank	0.000	15:42	
Control Test	0.079	15:42	
Air Blank	0.000	15:43	
Subject Sample	# 1	15:43	Subject Test Refused
Breath Volume	0.000	15:43	
Air Blank	0.000	15:43	
Control Test	0.079	15:44	
Air Blank	0.000	15:44	
Diagnostics Check	OK	15:44	

Data Download Date/Time: 07/28/2018 22:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAUGHN, DALE, E
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: COOPER, OFFICER
ARRESTING OFFICER AGENCY: Lake Placid PD
LAST AGENCY INSPECTION: 07/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: FERGUSON

DATE/TIME OF TEST: 07/29/2018 03:25:59

Diagnostics Check	OK	03:23	
Air Blank	0.000	03:23	
Control Test	0.078	03:23	
Air Blank	0.000	03:24	
Subject Sample	# 1	03:24	Subject Test Refused
Breath Volume	0.000	03:24	
Air Blank	0.000	03:25	
Control Test	0.078	03:25	
Air Blank	0.000	03:25	
Diagnostics Check	OK	03:25	

Data Download Date/Time: 08/26/2018 22:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,SUSAN,L
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: BOGUSKY,D
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:07

SUBJECT'S LAST NAME: VELA

DATE/TIME OF TEST: 07/29/2018 05:35:09

Diagnostics Check	OK	05:29
Air Blank	0.000	05:29
Control Test	0.078	05:29
Air Blank	0.000	05:30
Subject Sample # 1	0.189	05:30
Breath Volume	2.421	05:30
Air Blank	0.000	05:31
Air Blank	0.000	05:33
Subject Sample # 2	0.180	05:33
Breath Volume	2.285	05:33
Air Blank	0.000	05:34
Control Test	0.078	05:34
Air Blank	0.000	05:35
Diagnostics Check	OK	05:35

Data Download Date/Time: 08/26/2018 22:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAUGHN,DALE,E
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: CARVER,DEPUTY
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 07/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: PENA

DATE/TIME OF TEST: 07/30/2018 01:58:08

Diagnostics Check	OK	01:54	
Air Blank	0.000	01:55	
Control Test	0.078	01:55	
Air Blank	0.000	01:56	
Subject Sample	# 1	01:56	Subject Test Refused
Breath Volume	0.000	01:56	
Air Blank	0.000	01:57	
Control Test	0.078	01:57	
Air Blank	0.000	01:58	
Diagnostics Check	OK	01:58	

Data Download Date/Time: 08/26/2018 22:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,DUSTIN,J
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: RHODEN,DAVID
ARRESTING OFFICER AGENCY: Lake Placid PD
LAST AGENCY INSPECTION: 07/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: ORLANDI

DATE/TIME OF TEST: 07/30/2018 22:24:40

Diagnostics Check	OK	22:21	
Air Blank	0.000	22:22	
Control Test	0.079	22:22	
Air Blank	0.000	22:23	
Subject Sample	# 1	22:23	Subject Test Refused
Breath Volume	0.000	22:23	
Air Blank	0.000	22:23	
Control Test	0.079	22:24	
Air Blank	0.000	22:24	
Diagnostics Check	OK	22:24	

Data Download Date/Time: 08/26/2018 22:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

INSTRUMENT REGISTERED TO: DAVIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAINES,DAVID,K
BREATH TEST OPERATOR AGENCY: Davie PD
ARRESTING OFFICER: PUENTES,YENDY
ARRESTING OFFICER AGENCY: Davie PD
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 18:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:16

SUBJECT'S LAST NAME: DEL POZO

DATE/TIME OF TEST: 07/27/2018 18:47:32

Diagnostics Check	OK	18:40
Air Blank	0.000	18:41
Control Test	0.080	18:41
Air Blank	0.000	18:42
Subject Sample # 1	0.168	18:43
Breath Volume	1.847	18:43
Air Blank	0.000	18:43
Air Blank	0.000	18:45
Subject Sample # 2	0.154	18:46
Breath Volume	1.515	18:46
Air Blank	0.000	18:46
Control Test	0.079	18:46
Air Blank	0.000	18:47
Diagnostics Check	OK	18:47

Data Download Date/Time: 07/31/2018 18:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTLE,CORY,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SUTTLE,CORY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: ROSS

DATE/TIME OF TEST: 07/03/2018 03:58:56

Diagnostics Check	OK	03:52
Air Blank	0.000	03:53
Control Test	0.080	03:53
Air Blank	0.000	03:53
Subject Sample # 1	0.158	03:54
Breath Volume	2.835	03:54
Air Blank	0.000	03:55
Air Blank	0.000	03:56
Subject Sample # 2	0.153	03:57
Breath Volume	2.632	03:57
Air Blank	0.000	03:58
Control Test	0.079	03:58
Air Blank	0.000	03:58
Diagnostics Check	OK	03:58

Data Download Date/Time: 07/16/2018 08:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTLE,CORY,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SUTTLE,CORY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: IRVING

DATE/TIME OF TEST: 07/03/2018 22:22:33

Diagnostics Check	OK	22:16
Air Blank	0.000	22:17
Control Test	0.080	22:17
Air Blank	0.000	22:18
Subject Sample # 1	0.000	22:18
Breath Volume	2.207	22:18
Air Blank	0.000	22:18
Air Blank	0.000	22:20
Subject Sample # 2	0.000	22:21
Breath Volume	1.457	22:21
Air Blank	0.000	22:21
Control Test	0.080	22:21
Air Blank	0.000	22:22
Diagnostics Check	OK	22:22

Data Download Date/Time: 07/16/2018 08:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTLE,CORY,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SUTTLE,CORY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: JENKS

DATE/TIME OF TEST: 07/04/2018 03:21:10

Diagnostics Check	OK	03:15
Air Blank	0.000	03:15
Control Test	0.079	03:15
Air Blank	0.000	03:16
Subject Sample # 1	0.105	03:16
Breath Volume	2.660	03:16
Air Blank	0.000	03:17
Air Blank	0.000	03:19
Subject Sample # 2	0.095	03:19
Breath Volume	2.601	03:19
Air Blank	0.000	03:20
Control Test	0.078	03:20
Air Blank	0.000	03:21
Diagnostics Check	OK	03:21

Data Download Date/Time: 07/16/2018 08:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTLE,CORY,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SUTTLE,CORY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 07/06/2018 23:47:27

Diagnostics Check	OK	23:41
Air Blank	0.000	23:42
Control Test	0.079	23:42
Air Blank	0.000	23:42
Subject Sample # 1	0.000	23:43
Breath Volume	2.378	23:43
Air Blank	0.000	23:43
Air Blank	0.000	23:45
Subject Sample # 2	0.000	23:46
Breath Volume	2.492	23:46
Air Blank	0.000	23:46
Control Test	0.079	23:46
Air Blank	0.000	23:47
Diagnostics Check	OK	23:47

Data Download Date/Time: 07/16/2018 08:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTLE,CORY,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: STROM,ERIC
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 07/08/2018 02:18:21

Diagnostics Check	OK	02:12
Air Blank	0.000	02:12
Control Test	0.080	02:13
Air Blank	0.000	02:13
Subject Sample # 1	0.169	02:14
Breath Volume	2.484	02:14
Air Blank	0.000	02:14
Air Blank	0.000	02:16
Subject Sample # 2	0.166	02:16
Breath Volume	2.105	02:16
Air Blank	0.000	02:17
Control Test	0.079	02:17
Air Blank	0.000	02:18
Diagnostics Check	OK	02:18

Data Download Date/Time: 07/16/2018 08:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JASON,E
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: GOTHARD,RONALD
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 06/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 07/05/2018 03:16:26

Diagnostics Check	OK	03:13	
Air Blank	0.000	03:14	
Control Test	0.078	03:14	
Air Blank	0.000	03:14	
Subject Sample	# 1	03:15	Subject Test Refused
Breath Volume	0.000	03:15	
Air Blank	0.000	03:15	
Control Test	0.078	03:15	
Air Blank	0.000	03:16	
Diagnostics Check	OK	03:16	

Data Download Date/Time: 07/09/2018 07:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

INSTRUMENT REGISTERED TO: WINTER HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANTHAM,CODY,W
BREATH TEST OPERATOR AGENCY: Winter Haven PD
ARRESTING OFFICER: CORLEY,KRISTOPHER
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: WETLI

DATE/TIME OF TEST: 07/05/2018 00:00:55

Diagnostics Check	OK	23:55	
Air Blank	0.000	23:56	
Control Test	0.084	23:56	
Air Blank	0.000	23:57	
Subject Sample	# 1	23:59	Subject Test Refused
Breath Volume	0.000	23:59	
Air Blank	0.000	23:59	
Control Test	0.079	00:00	
Air Blank	0.000	00:00	
Diagnostics Check	OK	00:00	

Data Download Date/Time: 07/26/2018 16:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

INSTRUMENT REGISTERED TO: GULFPORT PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAROTTA,MICHAEL,S
BREATH TEST OPERATOR AGENCY: Gulfport PD
ARRESTING OFFICER: SMITH,CORY
ARRESTING OFFICER AGENCY: Gulfport PD
LAST AGENCY INSPECTION: 06/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: PINNOW

DATE/TIME OF TEST: 07/04/2018 22:39:11

Diagnostics Check	OK	22:32
Air Blank	0.000	22:33
Control Test	0.081	22:33
Air Blank	0.000	22:33
Subject Sample # 1	0.161	22:34
Breath Volume	2.320	22:34
Air Blank	0.000	22:35
Air Blank	0.000	22:37
Subject Sample # 2	0.160	22:37
Breath Volume	1.570	22:37
Air Blank	0.000	22:38
Control Test	0.079	22:38
Air Blank	0.000	22:39
Diagnostics Check	OK	22:39

Data Download Date/Time: 07/31/2018 15:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

INSTRUMENT REGISTERED TO: GULFPORT PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CARTER,TIMOTHY,R

BREATH TEST OPERATOR AGENCY: Gulfport PD

ARRESTING OFFICER: CARTER,TIMOTHY

ARRESTING OFFICER AGENCY: Gulfport PD

LAST AGENCY INSPECTION: 06/17/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17817080A2

DRY GAS EXPIRY DATE: 08/05/2019

VIOLATION CODE: DUI

ARREST TIME: 18:33

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:43

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 07/09/2018 19:13:42

Diagnostics Check	OK	19:03
Air Blank	0.000	19:03
Control Test	0.081	19:04
Air Blank	0.000	19:04
Subject Sample # 1	0.282	19:05
Breath Volume	1.691	19:05
Air Blank	0.000	19:05
Air Blank	0.000	19:07
Subject Sample # 2	0.259	19:08
Breath Volume	1.289	19:08
Air Blank	0.000	19:09
Air Blank	0.000	19:11
Subject Sample # 3	0.283	19:12
Breath Volume	3.195	19:12
Air Blank	0.000	19:12
Control Test	0.080	19:13
Air Blank	0.000	19:13
Diagnostics Check	OK	19:13

Data Download Date/Time: 07/31/2018 15:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001062

INSTRUMENT REGISTERED TO: TREASURE ISLAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORMAN,ZAKARIAH,S
BREATH TEST OPERATOR AGENCY: Treasure Island PD
ARRESTING OFFICER: DORMAN,ZAKARIAH
ARRESTING OFFICER AGENCY: Treasure Island PD
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: FOLSOM

DATE/TIME OF TEST: 07/07/2018 00:56:09

Diagnostics Check	OK	00:51	
Air Blank	0.000	00:52	
Control Test	0.079	00:52	
Air Blank	0.000	00:53	
Subject Sample	# 1	00:54	Subject Test Refused
Breath Volume	0.000	00:54	
Air Blank	0.000	00:55	
Control Test	0.080	00:55	
Air Blank	0.000	00:56	
Diagnostics Check	OK	00:56	

Data Download Date/Time: 07/27/2018 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001062

INSTRUMENT REGISTERED TO: TREASURE ISLAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORMAN,ZAKARIAH,S
BREATH TEST OPERATOR AGENCY: Treasure Island PD
ARRESTING OFFICER: HOLMS,MATT
ARRESTING OFFICER AGENCY: Indian Shores PD
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: ZIMNOCK

DATE/TIME OF TEST: 07/13/2018 03:37:50

Diagnostics Check	OK	03:31
Air Blank	0.000	03:31
Control Test	0.079	03:31
Air Blank	0.000	03:32
Subject Sample # 1	0.189	03:32
Breath Volume	3.304	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.193	03:36
Breath Volume	3.781	03:36
Air Blank	0.000	03:36
Control Test	0.079	03:37
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 07/27/2018 09:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001062

INSTRUMENT REGISTERED TO: TREASURE ISLAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HADDOCK,CLAY,S
BREATH TEST OPERATOR AGENCY: Treasure Island PD
ARRESTING OFFICER: MALANDER,CODY
ARRESTING OFFICER AGENCY: Treasure Island PD
LAST AGENCY INSPECTION: 07/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: CZERWONKA

DATE/TIME OF TEST: 07/31/2018 03:13:40

Diagnostics Check	OK	03:07
Air Blank	0.000	03:07
Control Test	0.077	03:07
Air Blank	0.000	03:08
Subject Sample # 1	0.252	03:09
Breath Volume	3.683	03:09
Air Blank	0.000	03:09
Air Blank	0.000	03:11
Subject Sample # 2	0.243	03:12
Breath Volume	3.152	03:12
Air Blank	0.000	03:12
Control Test	0.079	03:13
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

Data Download Date/Time: 08/30/2018 09:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,WILLIAM,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GREEN,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: PETRYNIEC

DATE/TIME OF TEST: 07/15/2018 02:37:59

Diagnostics Check	OK	02:31
Air Blank	0.000	02:31
Control Test	0.077	02:32
Air Blank	0.000	02:32
Subject Sample # 1	0.195	02:33
Breath Volume	1.906	02:33
Air Blank	0.000	02:33
Air Blank	0.000	02:35
Subject Sample # 2	0.200	02:36
Breath Volume	1.468	02:36
Air Blank	0.000	02:37
Control Test	0.075	02:37
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

Data Download Date/Time: 08/06/2018 22:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCONNELL,J,W
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: DELVALLE,M
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 07/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 15:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:11

SUBJECT'S LAST NAME: TONGEN

DATE/TIME OF TEST: 07/17/2018 16:54:53

Diagnostics Check	OK	16:46
Air Blank	0.000	16:46
Control Test	0.078	16:46
Air Blank	0.000	16:47
Subject Sample # 1	0.290	16:49
Breath Volume	1.597	16:49
Air Blank	0.000	16:49
Air Blank	0.000	16:51
Subject Sample # 2	0.286	16:53
Breath Volume	1.726	16:53
Air Blank	0.000	16:53
Control Test	0.078	16:54
Air Blank	0.000	16:54
Diagnostics Check	OK	16:54

Data Download Date/Time: 08/06/2018 22:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCONNELL,J,
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: GEE,T
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 07/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: JENNINGS

DATE/TIME OF TEST: 07/28/2018 04:15:37

Diagnostics Check	OK	04:04	
Air Blank	0.000	04:04	
Control Test	0.078	04:04	
Air Blank	0.000	04:05	
Subject Sample	# 1 0.174	04:08	Volume Not Met
Breath Volume	0.281	04:08	
Air Blank	0.000	04:09	
Air Blank	0.000	04:10	
Subject Sample	# 2 0.176	04:14	Volume Not Met
Breath Volume	0.039	04:14	
Air Blank	0.000	04:14	
Control Test	0.076	04:14	
Air Blank	0.000	04:15	
Diagnostics Check	OK	04:15	

Data Download Date/Time: 08/06/2018 22:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RONDOS,ROBERT,E
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: SCHELL,MATTHEW
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A3
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 07/08/2018 01:01:13

Diagnostics Check	OK	00:58	
Air Blank	0.000	00:58	
Control Test	0.078	00:59	
Air Blank	0.000	00:59	
Subject Sample	# 1	00:59	Subject Test Refused
Breath Volume	0.000	00:59	
Air Blank	0.000	01:00	
Control Test	0.078	01:00	
Air Blank	0.000	01:01	
Diagnostics Check	OK	01:01	

Data Download Date/Time: 07/30/2018 12:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRINIDAD,MICHAEL,E
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: SCHELL,M
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A3
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: GOREY

DATE/TIME OF TEST: 07/11/2018 03:23:09

Diagnostics Check	OK	03:16
Air Blank	0.000	03:17
Control Test	0.078	03:17
Air Blank	0.000	03:18
Subject Sample # 1	0.188	03:18
Breath Volume	1.382	03:18
Air Blank	0.000	03:19
Air Blank	0.000	03:21
Subject Sample # 2	0.201	03:21
Breath Volume	2.863	03:21
Air Blank	0.000	03:22
Control Test	0.077	03:22
Air Blank	0.000	03:23
Diagnostics Check	OK	03:23

Data Download Date/Time: 07/30/2018 12:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RONDOS,ROBERT,E
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: RONDOS,ROBERT
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A3
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:27

SUBJECT'S LAST NAME: CZYZ

DATE/TIME OF TEST: 07/17/2018 21:10:10

Diagnostics Check	OK	21:07	
Air Blank	0.000	21:07	
Control Test	0.078	21:07	
Air Blank	0.000	21:08	
Subject Sample	# 1	21:08	Subject Test Refused
Breath Volume	0.000	21:08	
Air Blank	0.000	21:09	
Control Test	0.077	21:09	
Air Blank	0.000	21:10	
Diagnostics Check	OK	21:10	

Data Download Date/Time: 07/30/2018 12:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,JACOB,R
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: RONDOS,ROBERT
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A3
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: AUMACK

DATE/TIME OF TEST: 07/22/2018 01:50:30

Diagnostics Check	OK	01:41
Air Blank	0.000	01:41
Control Test	0.079	01:41
Air Blank	0.000	01:42
Subject Sample # 1	0.210	01:45
Breath Volume	1.363	01:45
Air Blank	0.000	01:45
Air Blank	0.000	01:47
Subject Sample # 2	0.207	01:48
Breath Volume	1.164	01:48
Air Blank	0.000	01:49
Control Test	0.077	01:49
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

Data Download Date/Time: 07/30/2018 12:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: COOK,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: KILLIAN

DATE/TIME OF TEST: 07/01/2018 04:23:40

Diagnostics Check	OK	04:20	
Air Blank	0.000	04:20	
Control Test	0.081	04:21	
Air Blank	0.000	04:21	
Subject Sample	# 1	04:22	Subject Test Refused
Breath Volume	0.000	04:22	
Air Blank	0.000	04:22	
Control Test	0.081	04:23	
Air Blank	0.000	04:23	
Diagnostics Check	OK	04:23	

Data Download Date/Time: 07/26/2018 00:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHCOM, TIFFANY, J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MITCHELL, ROBERT
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 19:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:51

SUBJECT'S LAST NAME: WILLIAMSON

DATE/TIME OF TEST: 07/02/2018 20:17:42

Diagnostics Check	OK	20:11
Air Blank	0.000	20:12
Control Test	0.081	20:12
Air Blank	0.000	20:12
Subject Sample # 1	0.182	20:13
Breath Volume	2.035	20:13
Air Blank	0.000	20:13
Air Blank	0.000	20:15
Subject Sample # 2	0.176	20:16
Breath Volume	1.546	20:16
Air Blank	0.000	20:16
Control Test	0.081	20:17
Air Blank	0.000	20:17
Diagnostics Check	OK	20:17

Data Download Date/Time: 07/26/2018 00:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REFUSE, RONY,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: REFUSE, RONY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 13:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:30

SUBJECT'S LAST NAME: MCGLUE

DATE/TIME OF TEST: 07/03/2018 15:04:35

Diagnostics Check	OK	15:01	
Air Blank	0.000	15:01	
Control Test	0.081	15:02	
Air Blank	0.000	15:02	
Subject Sample	# 1	15:03	Subject Test Refused
Breath Volume	0.000	15:03	
Air Blank	0.000	15:03	
Control Test	0.081	15:03	
Air Blank	0.000	15:04	
Diagnostics Check	OK	15:04	

Data Download Date/Time: 07/26/2018 00:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTHONY, MICHAEL,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI, JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 07/03/2018 22:03:58

Diagnostics Check	OK	21:57
Air Blank	0.000	21:57
Control Test	0.082	21:58
Air Blank	0.000	21:58
Subject Sample # 1	0.000	21:59
Breath Volume	2.347	21:59
Air Blank	0.000	22:00
Air Blank	0.000	22:02
Subject Sample # 2	0.000	22:02
Breath Volume	2.613	22:02
Air Blank	0.000	22:03
Control Test	0.082	22:03
Air Blank	0.000	22:03
Diagnostics Check	OK	22:03

Data Download Date/Time: 07/26/2018 00:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: YUHASZ

DATE/TIME OF TEST: 07/04/2018 00:01:26

Diagnostics Check	OK	23:58	
Air Blank	0.000	23:59	
Control Test	0.082	23:59	
Air Blank	0.000	23:59	
Subject Sample	# 1	00:00	Subject Test Refused
Breath Volume	0.000	00:00	
Air Blank	0.000	00:00	
Control Test	0.082	00:00	
Air Blank	0.000	00:01	
Diagnostics Check	OK	00:01	

Data Download Date/Time: 07/26/2018 00:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTHONY,MICHAEL,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GAUTHIER,ERIN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: WASHINGTON

DATE/TIME OF TEST: 07/04/2018 01:20:23

Diagnostics Check	OK	01:11	
Air Blank	0.000	01:12	
Control Test	0.082	01:12	
Air Blank	0.000	01:13	
Subject Sample	# 1	01:16	No Sample Provided
Breath Volume	0.000	01:16	
Air Blank	0.000	01:16	
Air Blank	0.000	01:18	
Subject Sample	# 2	01:19	Subject Test Refused
Breath Volume	0.000	01:19	
Air Blank	0.000	01:19	
Control Test	0.082	01:19	
Air Blank	0.000	01:20	
Diagnostics Check	OK	01:20	

Data Download Date/Time: 07/26/2018 00:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: ATKINS JR

DATE/TIME OF TEST: 07/04/2018 04:35:00

Diagnostics Check	OK	04:29
Air Blank	0.000	04:29
Control Test	0.081	04:29
Air Blank	0.000	04:30
Subject Sample # 1	0.117	04:30
Breath Volume	2.453	04:30
Air Blank	0.000	04:31
Air Blank	0.000	04:33
Subject Sample # 2	0.121	04:33
Breath Volume	1.988	04:33
Air Blank	0.000	04:34
Control Test	0.081	04:34
Air Blank	0.000	04:34
Diagnostics Check	OK	04:35

Data Download Date/Time: 07/26/2018 00:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHCOM, TIFFANY, J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ROBINSON, JAMES
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 04:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:04

SUBJECT'S LAST NAME: DOTY

DATE/TIME OF TEST: 07/04/2018 05:27:52

Diagnostics Check	OK	05:24	
Air Blank	0.000	05:25	
Control Test	0.082	05:25	
Air Blank	0.000	05:26	
Subject Sample	# 1	05:26	Subject Test Refused
Breath Volume	0.000	05:26	
Air Blank	0.000	05:26	
Control Test	0.081	05:27	
Air Blank	0.000	05:27	
Diagnostics Check	OK	05:27	

Data Download Date/Time: 07/26/2018 00:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BOULTON,K
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: JIMENEZ

DATE/TIME OF TEST: 07/04/2018 19:24:02

Diagnostics Check	OK	19:21	
Air Blank	0.000	19:21	
Control Test	0.082	19:22	
Air Blank	0.000	19:22	
Subject Sample	# 1	19:22	Subject Test Refused
Breath Volume	0.000	19:22	
Air Blank	0.000	19:23	
Control Test	0.082	19:23	
Air Blank	0.000	19:23	
Diagnostics Check	OK	19:24	

Data Download Date/Time: 07/26/2018 00:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RANISZEWSKI,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:11

SUBJECT'S LAST NAME: AMSDEN

DATE/TIME OF TEST: 07/04/2018 22:41:33

Diagnostics Check	OK	22:35
Air Blank	0.000	22:35
Control Test	0.082	22:35
Air Blank	0.000	22:36
Subject Sample # 1	0.007	22:37
Breath Volume	1.152	22:37
Air Blank	0.000	22:37
Air Blank	0.000	22:39
Subject Sample # 2	0.000	22:40
Breath Volume	1.734	22:40
Air Blank	0.000	22:40
Control Test	0.082	22:40
Air Blank	0.000	22:41
Diagnostics Check	OK	22:41

Data Download Date/Time: 07/26/2018 00:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: LONG

DATE/TIME OF TEST: 07/05/2018 00:24:16

Diagnostics Check	OK	00:17
Air Blank	0.000	00:18
Control Test	0.082	00:18
Air Blank	0.000	00:18
Subject Sample # 1	0.236	00:19
Breath Volume	1.980	00:19
Air Blank	0.000	00:20
Air Blank	0.000	00:22
Subject Sample # 2	0.249	00:22
Breath Volume	1.234	00:22
Air Blank	0.000	00:23
Control Test	0.081	00:23
Air Blank	0.000	00:24
Diagnostics Check	OK	00:24

Data Download Date/Time: 07/26/2018 00:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: COROUTHERS

DATE/TIME OF TEST: 07/05/2018 01:43:20

Diagnostics Check	OK	01:37
Air Blank	0.000	01:37
Control Test	0.082	01:37
Air Blank	0.000	01:38
Subject Sample # 1	0.127	01:38
Breath Volume	2.148	01:38
Air Blank	0.000	01:39
Air Blank	0.000	01:41
Subject Sample # 2	0.129	01:41
Breath Volume	2.062	01:41
Air Blank	0.000	01:42
Control Test	0.081	01:42
Air Blank	0.000	01:43
Diagnostics Check	OK	01:43

Data Download Date/Time: 07/26/2018 00:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: SERHAN

DATE/TIME OF TEST: 07/05/2018 02:17:42

Diagnostics Check	OK	02:11
Air Blank	0.000	02:12
Control Test	0.081	02:12
Air Blank	0.000	02:13
Subject Sample # 1	0.046	02:13
Breath Volume	2.507	02:13
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample # 2	0.044	02:16
Breath Volume	2.550	02:16
Air Blank	0.000	02:16
Control Test	0.079	02:17
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

Data Download Date/Time: 07/26/2018 00:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: EASTRIDGE

DATE/TIME OF TEST: 07/06/2018 00:23:44

Diagnostics Check	OK	00:16
Air Blank	0.000	00:16
Control Test	0.082	00:17
Air Blank	0.000	00:17
Subject Sample # 1	0.240	00:18
Breath Volume	2.914	00:18
Air Blank	0.000	00:19
Air Blank	0.000	00:20
Subject Sample # 2	0.243	00:22
Breath Volume	3.652	00:22
Air Blank	0.000	00:22
Control Test	0.082	00:23
Air Blank	0.000	00:23
Diagnostics Check	OK	00:23

Data Download Date/Time: 07/26/2018 00:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: GREGUS-KIELEK

DATE/TIME OF TEST: 07/06/2018 01:47:57

Diagnostics Check	OK	01:45	
Air Blank	0.000	01:45	
Control Test	0.082	01:45	
Air Blank	0.000	01:46	
Subject Sample	# 1	01:46	Subject Test Refused
Breath Volume	0.000	01:46	
Air Blank	0.000	01:47	
Control Test	0.081	01:47	
Air Blank	0.000	01:47	
Diagnostics Check	OK	01:47	

Data Download Date/Time: 07/26/2018 00:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: ANGELINI

DATE/TIME OF TEST: 07/06/2018 04:53:19

Diagnostics Check	OK	04:50	
Air Blank	0.000	04:50	
Control Test	0.082	04:51	
Air Blank	0.000	04:51	
Subject Sample	# 1	04:51	Subject Test Refused
Breath Volume	0.000	04:51	
Air Blank	0.000	04:52	
Control Test	0.082	04:52	
Air Blank	0.000	04:53	
Diagnostics Check	OK	04:53	

Data Download Date/Time: 07/26/2018 00:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 07/07/2018 01:16:00

Diagnostics Check	OK	01:13	
Air Blank	0.000	01:13	
Control Test	0.082	01:13	
Air Blank	0.000	01:14	
Subject Sample	# 1	01:14	Subject Test Refused
Breath Volume	0.000	01:14	
Air Blank	0.000	01:15	
Control Test	0.082	01:15	
Air Blank	0.000	01:15	
Diagnostics Check	OK	01:16	

Data Download Date/Time: 07/26/2018 00:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: MESSINA

DATE/TIME OF TEST: 07/07/2018 02:05:49

Diagnostics Check	OK	02:02	
Air Blank	0.000	02:03	
Control Test	0.082	02:03	
Air Blank	0.000	02:04	
Subject Sample	# 1	02:04	Subject Test Refused
Breath Volume	0.000	02:04	
Air Blank	0.000	02:04	
Control Test	0.082	02:05	
Air Blank	0.000	02:05	
Diagnostics Check	OK	02:05	

Data Download Date/Time: 07/26/2018 00:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: CHOWDHURY

DATE/TIME OF TEST: 07/07/2018 03:52:46

Diagnostics Check	OK	03:47	
Air Blank	0.000	03:48	
Control Test	0.082	03:48	
Air Blank	0.000	03:48	
Subject Sample	# 1	03:51	Subject Test Refused
Breath Volume	0.000	03:51	
Air Blank	0.000	03:51	
Control Test	0.082	03:52	
Air Blank	0.000	03:52	
Diagnostics Check	OK	03:52	

Data Download Date/Time: 07/26/2018 00:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: HIGGINS

DATE/TIME OF TEST: 07/07/2018 04:22:49

Diagnostics Check	OK	04:15
Air Blank	0.000	04:15
Control Test	0.082	04:15
Air Blank	0.000	04:16
Subject Sample # 1	0.000	04:18
Breath Volume	1.703	04:18
Air Blank	0.000	04:18
Air Blank	0.000	04:20
Subject Sample # 2	0.000	04:21
Breath Volume	1.472	04:21
Air Blank	0.000	04:21
Control Test	0.082	04:22
Air Blank	0.000	04:22
Diagnostics Check	OK	04:22

Data Download Date/Time: 07/26/2018 00:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARTWICK,M,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FIVECOAT,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 11:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:48

SUBJECT'S LAST NAME: SEDLACK

DATE/TIME OF TEST: 07/07/2018 12:11:15

Diagnostics Check	OK	12:08	
Air Blank	0.000	12:08	
Control Test	0.082	12:09	
Air Blank	0.000	12:09	
Subject Sample	# 1	12:09	Subject Test Refused
Breath Volume	0.000	12:09	
Air Blank	0.000	12:10	
Control Test	0.082	12:10	
Air Blank	0.000	12:11	
Diagnostics Check	OK	12:11	

Data Download Date/Time: 07/26/2018 00:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTHONY,MICHAEL,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ANTHONY,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 18:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:28

SUBJECT'S LAST NAME: GRAHAM

DATE/TIME OF TEST: 07/07/2018 19:04:59

Diagnostics Check	OK	18:58
Air Blank	0.000	18:59
Control Test	0.082	18:59
Air Blank	0.000	19:00
Subject Sample # 1	0.148	19:00
Breath Volume	2.363	19:00
Air Blank	0.000	19:01
Air Blank	0.000	19:02
Subject Sample # 2	0.146	19:03
Breath Volume	2.285	19:03
Air Blank	0.000	19:04
Control Test	0.081	19:04
Air Blank	0.000	19:04
Diagnostics Check	OK	19:04

Data Download Date/Time: 07/26/2018 00:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: OBRIEN

DATE/TIME OF TEST: 07/08/2018 02:55:20

Diagnostics Check	OK	02:48
Air Blank	0.000	02:48
Control Test	0.082	02:49
Air Blank	0.000	02:49
Subject Sample # 1	0.090	02:50
Breath Volume	2.640	02:50
Air Blank	0.000	02:51
Air Blank	0.000	02:52
Subject Sample # 2	0.090	02:53
Breath Volume	2.871	02:53
Air Blank	0.000	02:54
Control Test	0.082	02:54
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

Data Download Date/Time: 07/26/2018 00:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HRYC,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: PALMA

DATE/TIME OF TEST: 07/08/2018 03:27:29

Diagnostics Check	OK	03:21
Air Blank	0.000	03:21
Control Test	0.082	03:21
Air Blank	0.000	03:22
Subject Sample # 1	0.186	03:23
Breath Volume	1.105	03:23
Air Blank	0.000	03:23
Air Blank	0.000	03:25
Subject Sample # 2	0.191	03:25
Breath Volume	1.339	03:25
Air Blank	0.000	03:26
Control Test	0.081	03:26
Air Blank	0.000	03:27
Diagnostics Check	OK	03:27

Data Download Date/Time: 07/26/2018 00:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: BALCH

DATE/TIME OF TEST: 07/08/2018 04:40:49

Diagnostics Check	OK	04:37	
Air Blank	0.000	04:38	
Control Test	0.082	04:38	
Air Blank	0.000	04:39	
Subject Sample	# 1	04:39	Subject Test Refused
Breath Volume	0.000	04:39	
Air Blank	0.000	04:39	
Control Test	0.082	04:40	
Air Blank	0.000	04:40	
Diagnostics Check	OK	04:40	

Data Download Date/Time: 07/26/2018 00:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARTWICK,M,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DARROW,M
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 13:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:38

SUBJECT'S LAST NAME: MALAVE

DATE/TIME OF TEST: 07/08/2018 15:04:54

Diagnostics Check	OK	14:58
Air Blank	0.000	14:59
Control Test	0.082	14:59
Air Blank	0.000	15:00
Subject Sample # 1	0.181	15:00
Breath Volume	1.613	15:00
Air Blank	0.000	15:01
Air Blank	0.000	15:03
Subject Sample # 2	0.179	15:03
Breath Volume	1.566	15:03
Air Blank	0.000	15:03
Control Test	0.081	15:04
Air Blank	0.000	15:04
Diagnostics Check	OK	15:04

Data Download Date/Time: 07/26/2018 00:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERSAUD,COMPTON,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PERSAUD,COMPTON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 07/09/2018 01:32:15

Diagnostics Check	OK	01:20	
Air Blank	0.000	01:21	
Control Test	0.082	01:21	
Air Blank	0.000	01:22	
Subject Sample	# 1	01:25	No Sample Provided
Breath Volume	0.000	01:25	
Air Blank	0.000	01:25	
Air Blank	0.000	01:27	
Subject Sample	# 2	01:30	No Sample Provided
Breath Volume	0.000	01:30	
Air Blank	0.000	01:31	
Control Test	0.082	01:31	
Air Blank	0.000	01:32	
Diagnostics Check	OK	01:32	

Data Download Date/Time: 07/26/2018 00:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,KYLE,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WEDE,BRENNEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: AGUILAR

DATE/TIME OF TEST: 07/09/2018 02:03:25

Diagnostics Check	OK	01:53
Air Blank	0.000	01:54
Control Test	0.083	01:54
Air Blank	0.000	01:54
Subject Sample # 1	0.295	01:55
Breath Volume	1.546	01:55
Air Blank	0.000	01:55
Air Blank	0.000	01:57
Subject Sample # 2	0.323	01:58
Breath Volume	2.207	01:58
Air Blank	0.000	01:59
Air Blank	0.000	02:01
Subject Sample # 3	0.304	02:01
Breath Volume	2.484	02:01
Air Blank	0.000	02:02
Control Test	0.081	02:02
Air Blank	0.000	02:03
Diagnostics Check	OK	02:03

Data Download Date/Time: 07/26/2018 00:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LEIB,M
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: LOGAN

DATE/TIME OF TEST: 07/12/2018 04:35:49

Diagnostics Check	OK	04:24	
Air Blank	0.000	04:24	
Control Test	0.082	04:25	
Air Blank	0.000	04:25	
Subject Sample	# 1 0.088	04:28	Volume Not Met
Breath Volume	0.867	04:28	
Air Blank	0.000	04:29	
Air Blank	0.000	04:31	
Subject Sample	# 2 0.076	04:34	Volume Not Met
Breath Volume	0.500	04:34	
Air Blank	0.000	04:34	
Control Test	0.082	04:35	
Air Blank	0.000	04:35	
Diagnostics Check	OK	04:35	

Data Download Date/Time: 07/26/2018 00:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: COOK,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: CURREY

DATE/TIME OF TEST: 07/14/2018 03:26:21

Diagnostics Check	OK	03:19
Air Blank	0.000	03:20
Control Test	0.082	03:20
Air Blank	0.000	03:21
Subject Sample # 1	0.041	03:21
Breath Volume	1.832	03:21
Air Blank	0.000	03:22
Air Blank	0.000	03:24
Subject Sample # 2	0.045	03:24
Breath Volume	1.773	03:24
Air Blank	0.000	03:25
Control Test	0.083	03:25
Air Blank	0.000	03:26
Diagnostics Check	OK	03:26

Data Download Date/Time: 07/26/2018 00:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 07/14/2018 03:36:28

Diagnostics Check	OK	03:30
Air Blank	0.000	03:30
Control Test	0.082	03:30
Air Blank	0.000	03:31
Subject Sample # 1	0.000	03:31
Breath Volume	1.917	03:31
Air Blank	0.000	03:32
Air Blank	0.000	03:34
Subject Sample # 2	0.000	03:35
Breath Volume	1.246	03:35
Air Blank	0.000	03:35
Control Test	0.082	03:35
Air Blank	0.000	03:36
Diagnostics Check	OK	03:36

Data Download Date/Time: 07/26/2018 00:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 05:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 07/14/2018 06:00:22

Diagnostics Check	OK	05:53
Air Blank	0.000	05:53
Control Test	0.082	05:54
Air Blank	0.000	05:54
Subject Sample # 1	0.086	05:55
Breath Volume	1.824	05:55
Air Blank	0.000	05:56
Air Blank	0.000	05:58
Subject Sample # 2	0.091	05:58
Breath Volume	2.628	05:58
Air Blank	0.000	05:59
Control Test	0.082	05:59
Air Blank	0.000	06:00
Diagnostics Check	OK	06:00

Data Download Date/Time: 07/26/2018 00:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBY,LARRY,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: JACOBY,LARRY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: PASCAZI

DATE/TIME OF TEST: 07/15/2018 00:29:04

Diagnostics Check	OK	00:20
Air Blank	0.000	00:21
Control Test	0.082	00:21
Air Blank	0.000	00:22
Subject Sample # 1	0.121	00:23
Breath Volume	1.335	00:23
Air Blank	0.000	00:24
Air Blank	0.000	00:26
Subject Sample # 2	0.118	00:27
Breath Volume	1.539	00:27
Air Blank	0.000	00:28
Control Test	0.082	00:28
Air Blank	0.000	00:28
Diagnostics Check	OK	00:29

Data Download Date/Time: 07/26/2018 00:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PUPKE,L
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: LOVETT

DATE/TIME OF TEST: 07/15/2018 02:16:07

Diagnostics Check	OK	02:13	
Air Blank	0.000	02:13	
Control Test	0.082	02:13	
Air Blank	0.000	02:14	
Subject Sample	# 1	02:14	Subject Test Refused
Breath Volume	0.000	02:14	
Air Blank	0.000	02:15	
Control Test	0.082	02:15	
Air Blank	0.000	02:16	
Diagnostics Check	OK	02:16	

Data Download Date/Time: 07/26/2018 00:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE,DEPUTY WILLIAM,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WILTSE,DEPUTY WILLIAM
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: MIXON

DATE/TIME OF TEST: 07/15/2018 03:57:20

Diagnostics Check	OK	03:54	
Air Blank	0.000	03:54	
Control Test	0.082	03:55	
Air Blank	0.000	03:55	
Subject Sample	# 1	03:55	Subject Test Refused
Breath Volume	0.000	03:55	
Air Blank	0.000	03:56	
Control Test	0.082	03:56	
Air Blank	0.000	03:57	
Diagnostics Check	OK	03:57	

Data Download Date/Time: 07/26/2018 00:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHCOM, TIFFANY, J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LEIB, MELANIE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:12

SUBJECT'S LAST NAME: HEANEY

DATE/TIME OF TEST: 07/16/2018 21:39:00

Diagnostics Check	OK	21:32
Air Blank	0.000	21:33
Control Test	0.077	21:33
Air Blank	0.000	21:33
Subject Sample # 1	0.194	21:34
Breath Volume	1.292	21:34
Air Blank	0.000	21:35
Air Blank	0.000	21:36
Subject Sample # 2	0.186	21:37
Breath Volume	1.753	21:37
Air Blank	0.000	21:38
Control Test	0.080	21:38
Air Blank	0.000	21:38
Diagnostics Check	OK	21:39

Data Download Date/Time: 07/26/2018 00:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORNING,PETER,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DENHARD,JUSTIN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 16:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:39

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 07/17/2018 18:04:12

Diagnostics Check	OK	18:01	
Air Blank	0.000	18:01	
Control Test	0.081	18:01	
Air Blank	0.000	18:02	
Subject Sample	# 1	18:02	Subject Test Refused
Breath Volume	0.000	18:02	
Air Blank	0.000	18:03	
Control Test	0.081	18:03	
Air Blank	0.000	18:04	
Diagnostics Check	OK	18:04	

Data Download Date/Time: 07/26/2018 00:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHCOM, TIFFANY, J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ASHCOM, TIFFANY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: ARCHER

DATE/TIME OF TEST: 07/18/2018 01:49:23

Diagnostics Check	OK	01:43
Air Blank	0.000	01:43
Control Test	0.081	01:44
Air Blank	0.000	01:44
Subject Sample # 1	0.077	01:45
Breath Volume	2.800	01:45
Air Blank	0.000	01:45
Air Blank	0.000	01:47
Subject Sample # 2	0.077	01:47
Breath Volume	3.328	01:47
Air Blank	0.000	01:48
Control Test	0.079	01:48
Air Blank	0.000	01:49
Diagnostics Check	OK	01:49

Data Download Date/Time: 07/26/2018 00:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: FORTE

DATE/TIME OF TEST: 07/19/2018 03:08:09

Diagnostics Check	OK	03:05	
Air Blank	0.000	03:05	
Control Test	0.081	03:06	
Air Blank	0.000	03:06	
Subject Sample	# 1	03:06	Subject Test Refused
Breath Volume	0.000	03:06	
Air Blank	0.000	03:07	
Control Test	0.081	03:07	
Air Blank	0.000	03:08	
Diagnostics Check	OK	03:08	

Data Download Date/Time: 07/26/2018 00:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MULLINS,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: DANIELS

DATE/TIME OF TEST: 07/20/2018 01:08:22

Diagnostics Check	OK	01:05	
Air Blank	0.000	01:06	
Control Test	0.081	01:06	
Air Blank	0.000	01:06	
Subject Sample	# 1	01:07	Subject Test Refused
Breath Volume	0.000	01:07	
Air Blank	0.000	01:07	
Control Test	0.080	01:07	
Air Blank	0.000	01:08	
Diagnostics Check	OK	01:08	

Data Download Date/Time: 07/26/2018 00:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: BONK

DATE/TIME OF TEST: 07/20/2018 03:01:10

Diagnostics Check	OK	02:58	
Air Blank	0.000	02:58	
Control Test	0.081	02:59	
Air Blank	0.000	02:59	
Subject Sample	# 1	02:59	Subject Test Refused
Breath Volume	0.000	02:59	
Air Blank	0.000	03:00	
Control Test	0.082	03:00	
Air Blank	0.000	03:01	
Diagnostics Check	OK	03:01	

Data Download Date/Time: 07/26/2018 00:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: PAULK

DATE/TIME OF TEST: 07/21/2018 03:56:42

Diagnostics Check	OK	03:52	
Air Blank	0.000	03:53	
Control Test	0.082	03:53	
Air Blank	0.000	03:53	
Subject Sample	# 1	03:55	Subject Test Refused
Breath Volume	0.000	03:55	
Air Blank	0.000	03:55	
Control Test	0.082	03:56	
Air Blank	0.000	03:56	
Diagnostics Check	OK	03:56	

Data Download Date/Time: 07/26/2018 00:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARTWICK,MICHAEL,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MCTAGGART,T
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 13:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:55

SUBJECT'S LAST NAME: THAYER

DATE/TIME OF TEST: 07/21/2018 14:22:25

Diagnostics Check	OK	14:15
Air Blank	0.000	14:15
Control Test	0.081	14:15
Air Blank	0.000	14:16
Subject Sample # 1	0.296	14:17
Breath Volume	1.171	14:17
Air Blank	0.000	14:17
Air Blank	0.000	14:19
Subject Sample # 2	0.288	14:20
Breath Volume	1.480	14:20
Air Blank	0.000	14:21
Control Test	0.080	14:21
Air Blank	0.000	14:22
Diagnostics Check	OK	14:22

Data Download Date/Time: 07/26/2018 00:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMIN,N
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 18:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 07/21/2018 19:15:55

Diagnostics Check	OK	19:04	
Air Blank	0.000	19:04	
Control Test	0.081	19:05	
Air Blank	0.000	19:05	
Subject Sample	# 1 0.314	19:08	Volume Not Met
Breath Volume	0.519	19:08	
Air Blank	0.000	19:09	
Air Blank	0.000	19:11	
Subject Sample	# 2 0.275	19:14	Volume Not Met
Breath Volume	0.101	19:14	
Air Blank	0.000	19:14	
Control Test	0.078	19:15	
Air Blank	0.000	19:15	
Diagnostics Check	OK	19:15	

Data Download Date/Time: 07/26/2018 00:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: KICO

DATE/TIME OF TEST: 07/21/2018 22:59:29

Diagnostics Check	OK	22:52
Air Blank	0.000	22:53
Control Test	0.081	22:53
Air Blank	0.000	22:53
Subject Sample # 1	0.000	22:55
Breath Volume	1.496	22:55
Air Blank	0.000	22:55
Air Blank	0.000	22:57
Subject Sample # 2	0.000	22:58
Breath Volume	1.519	22:58
Air Blank	0.000	22:58
Control Test	0.081	22:58
Air Blank	0.000	22:59
Diagnostics Check	OK	22:59

Data Download Date/Time: 07/26/2018 00:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORNING,PETER,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DUNNING,RYAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 16:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: MCLAUGHLIN

DATE/TIME OF TEST: 07/22/2018 17:43:48

Diagnostics Check	OK	17:35	
Air Blank	0.000	17:35	
Control Test	0.081	17:36	
Air Blank	0.000	17:36	
Subject Sample	# 1	17:39	No Sample Provided
Breath Volume	0.000	17:39	
Air Blank	0.000	17:40	
Air Blank	0.000	17:42	
Subject Sample	# 2	17:42	Subject Test Refused
Breath Volume	0.000	17:42	
Air Blank	0.000	17:42	
Control Test	0.081	17:43	
Air Blank	0.000	17:43	
Diagnostics Check	OK	17:43	

Data Download Date/Time: 07/26/2018 00:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 21:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: JASMIN

DATE/TIME OF TEST: 07/22/2018 22:18:41

Diagnostics Check	OK	22:12
Air Blank	0.000	22:13
Control Test	0.081	22:13
Air Blank	0.000	22:13
Subject Sample # 1	0.066	22:14
Breath Volume	1.593	22:14
Air Blank	0.000	22:14
Air Blank	0.000	22:16
Subject Sample # 2	0.073	22:17
Breath Volume	2.109	22:17
Air Blank	0.000	22:17
Control Test	0.080	22:18
Air Blank	0.000	22:18
Diagnostics Check	OK	22:18

Data Download Date/Time: 07/26/2018 00:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBY,LARRY,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CORSER,WILLIAM
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 17:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:18

SUBJECT'S LAST NAME: SCULLY

DATE/TIME OF TEST: 07/23/2018 18:46:47

Diagnostics Check	OK	18:38
Air Blank	0.000	18:38
Control Test	0.082	18:39
Air Blank	0.000	18:39
Subject Sample # 1	0.135	18:42
Breath Volume	3.855	18:42
Air Blank	0.000	18:42
Air Blank	0.000	18:44
Subject Sample # 2	0.127	18:45
Breath Volume	2.382	18:45
Air Blank	0.000	18:45
Control Test	0.079	18:46
Air Blank	0.000	18:46
Diagnostics Check	OK	18:46

Data Download Date/Time: 07/26/2018 00:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,NELSON,X
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: COOK,CHRISTOPHER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: MITROVIC

DATE/TIME OF TEST: 07/25/2018 00:46:41

Diagnostics Check	OK	00:40
Air Blank	0.000	00:41
Control Test	0.082	00:41
Air Blank	0.000	00:41
Subject Sample # 1	0.000	00:42
Breath Volume	3.539	00:42
Air Blank	0.000	00:42
Air Blank	0.000	00:44
Subject Sample # 2	0.000	00:45
Breath Volume	2.945	00:45
Air Blank	0.000	00:45
Control Test	0.081	00:46
Air Blank	0.000	00:46
Diagnostics Check	OK	00:46

Data Download Date/Time: 07/26/2018 00:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: WOLSTEIN

DATE/TIME OF TEST: 07/26/2018 03:20:41

Diagnostics Check	OK	03:09	
Air Blank	0.000	03:09	
Control Test	0.081	03:09	
Air Blank	0.000	03:10	
Subject Sample	# 1	03:13	No Sample Provided
Breath Volume	0.000	03:13	
Air Blank	0.000	03:13	
Air Blank	0.000	03:15	
Subject Sample	# 2	0.202	03:19
Breath Volume	0.894	03:19	Volume Not Met
Air Blank	0.000	03:19	
Control Test	0.082	03:20	
Air Blank	0.000	03:20	
Diagnostics Check	OK	03:20	

Data Download Date/Time: 08/02/2018 23:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: WOLSTEIN

DATE/TIME OF TEST: 07/26/2018 03:35:55

Diagnostics Check	OK	03:24	
Air Blank	0.000	03:24	
Control Test	0.081	03:25	
Air Blank	0.000	03:25	
Subject Sample	# 1 0.211	03:28	Volume Not Met
Breath Volume	0.539	03:28	
Air Blank	0.000	03:29	
Air Blank	0.000	03:31	
Subject Sample	# 2 0.182	03:34	Volume Not Met
Breath Volume	0.371	03:34	
Air Blank	0.000	03:35	
Control Test	0.079	03:35	
Air Blank	0.000	03:35	
Diagnostics Check	OK	03:35	

Data Download Date/Time: 08/02/2018 23:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FOSTER,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: HIGSON

DATE/TIME OF TEST: 07/26/2018 04:27:23

Diagnostics Check	OK	04:21
Air Blank	0.000	04:21
Control Test	0.081	04:21
Air Blank	0.000	04:22
Subject Sample # 1	0.180	04:22
Breath Volume	1.257	04:22
Air Blank	0.000	04:23
Air Blank	0.000	04:25
Subject Sample # 2	0.187	04:25
Breath Volume	1.148	04:25
Air Blank	0.000	04:26
Control Test	0.080	04:26
Air Blank	0.000	04:27
Diagnostics Check	OK	04:27

Data Download Date/Time: 08/02/2018 23:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHCOM, TIFFANY, J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ASHCOM, TIFFANY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 19:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:41

SUBJECT'S LAST NAME: CASTELLO

DATE/TIME OF TEST: 07/26/2018 21:06:28

Diagnostics Check	OK	21:03	
Air Blank	0.000	21:04	
Control Test	0.081	21:04	
Air Blank	0.000	21:04	
Subject Sample	# 1	21:05	Subject Test Refused
Breath Volume	0.000	21:05	
Air Blank	0.000	21:05	
Control Test	0.081	21:05	
Air Blank	0.000	21:06	
Diagnostics Check	OK	21:06	

Data Download Date/Time: 08/02/2018 23:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTHONY, MICHAEL,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FOSTER, ALEXANDER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: MATTHEWS

DATE/TIME OF TEST: 07/26/2018 23:59:06

Diagnostics Check	OK	23:55	
Air Blank	0.000	23:56	
Control Test	0.081	23:56	
Air Blank	0.000	23:57	
Subject Sample	# 1	23:57	Subject Test Refused
Breath Volume	0.000	23:57	
Air Blank	0.000	23:58	
Control Test	0.080	23:58	
Air Blank	0.000	23:59	
Diagnostics Check	OK	23:59	

Data Download Date/Time: 08/02/2018 23:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTHONY, MICHAEL,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RANISZEWSKI, ANDREW
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: HOLT

DATE/TIME OF TEST: 07/27/2018 02:10:13

Diagnostics Check	OK	02:04	
Air Blank	0.000	02:05	
Control Test	0.081	02:05	
Air Blank	0.000	02:06	
Subject Sample	# 1	02:08	Subject Test Refused
Breath Volume	0.000	02:08	
Air Blank	0.000	02:09	
Control Test	0.080	02:09	
Air Blank	0.000	02:10	
Diagnostics Check	OK	02:10	

Data Download Date/Time: 08/02/2018 23:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTHONY, MICHAEL,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE, LIZ
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: JEZEK

DATE/TIME OF TEST: 07/27/2018 03:20:12

Diagnostics Check	OK	03:13
Air Blank	0.000	03:13
Control Test	0.080	03:14
Air Blank	0.000	03:14
Subject Sample # 1	0.183	03:15
Breath Volume	1.398	03:15
Air Blank	0.000	03:15
Air Blank	0.000	03:17
Subject Sample # 2	0.181	03:18
Breath Volume	2.085	03:18
Air Blank	0.000	03:19
Control Test	0.080	03:19
Air Blank	0.000	03:20
Diagnostics Check	OK	03:20

Data Download Date/Time: 08/02/2018 23:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTHONY,MICHAEL,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FRASER,CHRISTOPHER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: RONACHER

DATE/TIME OF TEST: 07/28/2018 00:15:31

Diagnostics Check	OK	00:09	
Air Blank	0.000	00:10	
Control Test	0.081	00:10	
Air Blank	0.000	00:11	
Subject Sample	# 1	00:14	Subject Test Refused
Breath Volume	0.000	00:14	
Air Blank	0.000	00:14	
Control Test	0.081	00:14	
Air Blank	0.000	00:15	
Diagnostics Check	OK	00:15	

Data Download Date/Time: 08/02/2018 23:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERSAUD,COMPTON,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PERSAUD,COMPTON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: PENNING

DATE/TIME OF TEST: 07/28/2018 02:22:51

Diagnostics Check	OK	02:11	
Air Blank	0.000	02:11	
Control Test	0.081	02:12	
Air Blank	0.000	02:12	
Subject Sample	# 1	02:15	No Sample Provided
Breath Volume	0.000	02:15	
Air Blank	0.000	02:16	
Air Blank	0.000	02:18	
Subject Sample	# 2	02:21	No Sample Provided
Breath Volume	0.000	02:21	
Air Blank	0.000	02:21	
Control Test	0.081	02:22	
Air Blank	0.000	02:22	
Diagnostics Check	OK	02:22	

Data Download Date/Time: 08/02/2018 23:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBY,LARRY,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: JACOBY,LARRY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: SHAW

DATE/TIME OF TEST: 07/28/2018 04:26:34

Diagnostics Check	OK	04:20
Air Blank	0.000	04:20
Control Test	0.080	04:20
Air Blank	0.000	04:21
Subject Sample # 1	0.112	04:22
Breath Volume	3.121	04:22
Air Blank	0.000	04:22
Air Blank	0.000	04:24
Subject Sample # 2	0.115	04:25
Breath Volume	3.917	04:25
Air Blank	0.000	04:25
Control Test	0.080	04:25
Air Blank	0.000	04:26
Diagnostics Check	OK	04:26

Data Download Date/Time: 08/02/2018 23:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTHONY, MICHAEL,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MULLINS, JOHN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: MALDONADO

DATE/TIME OF TEST: 07/28/2018 05:01:22

Diagnostics Check	OK	04:53
Air Blank	0.000	04:54
Control Test	0.080	04:54
Air Blank	0.000	04:55
Subject Sample # 1	0.106	04:56
Breath Volume	2.167	04:56
Air Blank	0.000	04:57
Air Blank	0.000	04:59
Subject Sample # 2	0.103	04:59
Breath Volume	2.125	04:59
Air Blank	0.000	05:00
Control Test	0.080	05:00
Air Blank	0.000	05:01
Diagnostics Check	OK	05:01

Data Download Date/Time: 08/02/2018 23:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE,DEPUTY WILLIAM,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FRASER,DEPUTY C.
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: ROUSSEAU

DATE/TIME OF TEST: 07/29/2018 01:07:19

Diagnostics Check	OK	01:00
Air Blank	0.000	01:00
Control Test	0.081	01:01
Air Blank	0.000	01:01
Subject Sample # 1	0.254	01:02
Breath Volume	1.781	01:02
Air Blank	0.000	01:03
Air Blank	0.000	01:05
Subject Sample # 2	0.252	01:05
Breath Volume	1.601	01:05
Air Blank	0.000	01:06
Control Test	0.079	01:06
Air Blank	0.000	01:07
Diagnostics Check	OK	01:07

Data Download Date/Time: 08/02/2018 23:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE,DEPUTY WILLIAM,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: COOK,DEPUTY C.
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 19:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: FREDERICK

DATE/TIME OF TEST: 07/29/2018 20:47:05

Diagnostics Check	OK	20:32	
Air Blank	0.000	20:33	
Control Test	0.081	20:33	
Air Blank	0.000	20:34	
Subject Sample	# 1 0.174	20:37	Volume Not Met
Breath Volume	1.183	20:37	
Air Blank	0.000	20:37	
Air Blank	0.000	20:39	
Subject Sample	# 2 0.187	20:41	
Breath Volume	2.207	20:41	
Air Blank	0.000	20:42	
Air Blank	0.000	20:44	
Subject Sample	# 3 0.181	20:45	
Breath Volume	1.535	20:45	
Air Blank	0.000	20:46	
Control Test	0.078	20:46	
Air Blank	0.000	20:47	
Diagnostics Check	OK	20:47	

Data Download Date/Time: 08/02/2018 23:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE,DEPUTY WILLIAM,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FRASER,DEPUTY C.
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: FREED

DATE/TIME OF TEST: 07/29/2018 23:44:17

Diagnostics Check	OK	23:32	
Air Blank	0.000	23:33	
Control Test	0.081	23:33	
Air Blank	0.000	23:34	
Subject Sample	# 1	23:37	No Sample Provided
Breath Volume	0.000	23:37	
Air Blank	0.000	23:37	
Air Blank	0.000	23:39	
Subject Sample	# 2	23:42	No Sample Provided
Breath Volume	0.000	23:42	
Air Blank	0.000	23:43	
Control Test	0.081	23:43	
Air Blank	0.000	23:44	
Diagnostics Check	OK	23:44	

Data Download Date/Time: 08/02/2018 23:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE,DEPUTY WILLIAM,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FRASER,DEPUTY C.
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: FREED

DATE/TIME OF TEST: 07/29/2018 23:57:47

Diagnostics Check	OK	23:50
Air Blank	0.000	23:51
Control Test	0.079	23:51
Air Blank	0.000	23:52
Subject Sample # 1	0.215	23:52
Breath Volume	1.710	23:52
Air Blank	0.000	23:53
Air Blank	0.000	23:55
Subject Sample # 2	0.214	23:56
Breath Volume	2.910	23:56
Air Blank	0.000	23:56
Control Test	0.080	23:57
Air Blank	0.000	23:57
Diagnostics Check	OK	23:57

Data Download Date/Time: 08/02/2018 23:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROSS,MARK,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GEOHEGAN,LEAH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: CARDENAS ALVAREZ

DATE/TIME OF TEST: 07/30/2018 03:44:35

Diagnostics Check	OK	03:38
Air Blank	0.000	03:38
Control Test	0.080	03:39
Air Blank	0.000	03:39
Subject Sample # 1	0.000	03:40
Breath Volume	3.062	03:40
Air Blank	0.000	03:40
Air Blank	0.000	03:42
Subject Sample # 2	0.000	03:43
Breath Volume	2.074	03:43
Air Blank	0.000	03:43
Control Test	0.080	03:43
Air Blank	0.000	03:44
Diagnostics Check	OK	03:44

Data Download Date/Time: 08/02/2018 23:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: STRZALKOWSKI,SEAN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: DODEK

DATE/TIME OF TEST: 07/03/2018 03:21:56

Diagnostics Check	OK	03:14
Air Blank	0.000	03:14
Control Test	0.079	03:15
Air Blank	0.000	03:15
Subject Sample # 1	0.159	03:16
Breath Volume	1.113	03:16
Air Blank	0.000	03:16
Air Blank	0.000	03:18
Subject Sample # 2	0.159	03:20
Breath Volume	1.289	03:20
Air Blank	0.000	03:21
Control Test	0.079	03:21
Air Blank	0.000	03:21
Diagnostics Check	OK	03:21

Data Download Date/Time: 07/13/2018 11:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: STRZALKOWSKI,SEAN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: WALTON

DATE/TIME OF TEST: 07/04/2018 01:21:21

Diagnostics Check	OK	01:14
Air Blank	0.000	01:14
Control Test	0.079	01:14
Air Blank	0.000	01:15
Subject Sample # 1	0.151	01:16
Breath Volume	1.304	01:16
Air Blank	0.000	01:17
Air Blank	0.000	01:18
Subject Sample # 2	0.148	01:19
Breath Volume	1.222	01:19
Air Blank	0.000	01:20
Control Test	0.078	01:20
Air Blank	0.000	01:21
Diagnostics Check	OK	01:21

Data Download Date/Time: 07/13/2018 11:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MARINO,GREG
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 04:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:01

SUBJECT'S LAST NAME: TRUJILLO

DATE/TIME OF TEST: 07/06/2018 05:33:44

Diagnostics Check	OK	05:22	
Air Blank	0.000	05:22	
Control Test	0.079	05:22	
Air Blank	0.000	05:23	
Subject Sample	# 1 0.136	05:26	Volume Not Met
Breath Volume	0.082	05:26	
Air Blank	0.000	05:27	
Air Blank	0.000	05:29	
Subject Sample	# 2 0.122	05:32	Volume Not Met
Breath Volume	0.000	05:32	
Air Blank	0.000	05:32	
Control Test	0.078	05:33	
Air Blank	0.000	05:33	
Diagnostics Check	OK	05:33	

Data Download Date/Time: 07/13/2018 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: GIBOSSE

DATE/TIME OF TEST: 07/07/2018 04:40:17

Diagnostics Check	OK	04:33
Air Blank	0.000	04:34
Control Test	0.078	04:34
Air Blank	0.000	04:35
Subject Sample # 1	0.101	04:35
Breath Volume	1.765	04:35
Air Blank	0.000	04:36
Air Blank	0.000	04:38
Subject Sample # 2	0.099	04:38
Breath Volume	2.421	04:38
Air Blank	0.000	04:39
Control Test	0.078	04:39
Air Blank	0.000	04:40
Diagnostics Check	OK	04:40

Data Download Date/Time: 07/13/2018 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ICOBELLI,R
ARRESTING OFFICER AGENCY: Hallandale Beach PD
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 07/08/2018 02:34:11

Diagnostics Check	OK	02:27
Air Blank	0.000	02:27
Control Test	0.079	02:27
Air Blank	0.000	02:28
Subject Sample # 1	0.118	02:28
Breath Volume	1.222	02:28
Air Blank	0.000	02:29
Air Blank	0.000	02:31
Subject Sample # 2	0.110	02:32
Breath Volume	2.035	02:32
Air Blank	0.000	02:33
Control Test	0.079	02:33
Air Blank	0.000	02:34
Diagnostics Check	OK	02:34

Data Download Date/Time: 07/13/2018 11:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SOUZA,MAYKOLO
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: MASON

DATE/TIME OF TEST: 07/13/2018 00:30:55

Diagnostics Check	OK	00:24
Air Blank	0.000	00:25
Control Test	0.079	00:25
Air Blank	0.000	00:25
Subject Sample # 1	0.204	00:26
Breath Volume	2.234	00:26
Air Blank	0.000	00:27
Air Blank	0.000	00:28
Subject Sample # 2	0.195	00:29
Breath Volume	1.734	00:29
Air Blank	0.000	00:30
Control Test	0.079	00:30
Air Blank	0.000	00:30
Diagnostics Check	OK	00:30

Data Download Date/Time: 07/13/2018 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: CASTROLUX

DATE/TIME OF TEST: 07/13/2018 02:11:45

Diagnostics Check	OK	02:05
Air Blank	0.000	02:05
Control Test	0.079	02:06
Air Blank	0.000	02:06
Subject Sample # 1	0.218	02:07
Breath Volume	3.050	02:07
Air Blank	0.000	02:07
Air Blank	0.000	02:09
Subject Sample # 2	0.222	02:10
Breath Volume	3.437	02:10
Air Blank	0.000	02:10
Control Test	0.079	02:11
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

Data Download Date/Time: 07/13/2018 11:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HURT,JONATHAN,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: HURT,JONATHAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 07/04/2018 01:56:38

Diagnostics Check	OK	01:49
Air Blank	0.000	01:50
Control Test	0.079	01:50
Air Blank	0.000	01:51
Subject Sample # 1	0.000	01:52
Breath Volume	4.117	01:52
Air Blank	0.000	01:52
Air Blank	0.000	01:54
Subject Sample # 2	0.000	01:55
Breath Volume	4.023	01:55
Air Blank	0.000	01:55
Control Test	0.079	01:56
Air Blank	0.000	01:56
Diagnostics Check	OK	01:56

Data Download Date/Time: 07/20/2018 20:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEGERSMITH,DANIEL,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,DANIEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: THOME

DATE/TIME OF TEST: 07/05/2018 20:47:58

Diagnostics Check	OK	20:41
Air Blank	0.000	20:41
Control Test	0.080	20:41
Air Blank	0.000	20:42
Subject Sample # 1	0.000	20:43
Breath Volume	2.269	20:43
Air Blank	0.000	20:43
Air Blank	0.000	20:45
Subject Sample # 2	0.000	20:46
Breath Volume	2.187	20:46
Air Blank	0.000	20:47
Control Test	0.079	20:47
Air Blank	0.000	20:47
Diagnostics Check	OK	20:47

Data Download Date/Time: 07/20/2018 20:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEGERSMITH,DANIEL,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,DANIEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 07/13/2018 00:36:10

Diagnostics Check	OK	00:28
Air Blank	0.000	00:29
Control Test	0.079	00:29
Air Blank	0.000	00:30
Subject Sample # 1	0.000	00:31
Breath Volume	1.339	00:31
Air Blank	0.000	00:31
Air Blank	0.000	00:33
Subject Sample # 2	0.000	00:34
Breath Volume	2.085	00:34
Air Blank	0.000	00:35
Control Test	0.081	00:35
Air Blank	0.000	00:36
Diagnostics Check	OK	00:36

Data Download Date/Time: 07/20/2018 20:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HURT,JONATHAN,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: HURT,JONATHAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: ERICKSON

DATE/TIME OF TEST: 07/14/2018 02:47:33

Diagnostics Check	OK	02:40
Air Blank	0.000	02:41
Control Test	0.079	02:41
Air Blank	0.000	02:42
Subject Sample # 1	0.184	02:43
Breath Volume	2.546	02:43
Air Blank	0.000	02:43
Air Blank	0.000	02:45
Subject Sample # 2	0.180	02:45
Breath Volume	1.210	02:45
Air Blank	0.000	02:46
Control Test	0.078	02:46
Air Blank	0.000	02:47
Diagnostics Check	OK	02:47

Data Download Date/Time: 07/20/2018 20:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MICHEO,FRANCISCO,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: MICHEO,FRANCISCO
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 07/15/2018 19:26:35

Diagnostics Check	OK	19:19
Air Blank	0.000	19:19
Control Test	0.079	19:19
Air Blank	0.000	19:20
Subject Sample # 1	0.133	19:21
Breath Volume	1.957	19:21
Air Blank	0.000	19:22
Air Blank	0.000	19:24
Subject Sample # 2	0.129	19:24
Breath Volume	2.828	19:24
Air Blank	0.000	19:25
Control Test	0.078	19:25
Air Blank	0.000	19:26
Diagnostics Check	OK	19:26

Data Download Date/Time: 07/20/2018 20:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEGERSMITH,DANIEL,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,DANIEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: HAYES

DATE/TIME OF TEST: 07/17/2018 01:07:45

Diagnostics Check	OK	00:56	
Air Blank	0.000	00:56	
Control Test	0.079	00:57	
Air Blank	0.000	00:57	
Subject Sample	# 1	01:00	No Sample Provided
Breath Volume	0.000	01:00	
Air Blank	0.000	01:01	
Air Blank	0.000	01:03	
Subject Sample	# 2	01:06	No Sample Provided
Breath Volume	0.000	01:06	
Air Blank	0.000	01:06	
Control Test	0.080	01:07	
Air Blank	0.000	01:07	
Diagnostics Check	OK	01:07	

Data Download Date/Time: 07/20/2018 20:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOXIE,BRIAN,P
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: WILLIAMS,JEFF
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: STORR

DATE/TIME OF TEST: 07/18/2018 23:30:18

Diagnostics Check	OK	23:23
Air Blank	0.000	23:23
Control Test	0.079	23:23
Air Blank	0.000	23:24
Subject Sample # 1	0.147	23:25
Breath Volume	1.953	23:25
Air Blank	0.000	23:26
Air Blank	0.000	23:28
Subject Sample # 2	0.145	23:28
Breath Volume	2.058	23:28
Air Blank	0.000	23:29
Control Test	0.079	23:29
Air Blank	0.000	23:30
Diagnostics Check	OK	23:30

Data Download Date/Time: 07/20/2018 20:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,CHAD,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: HURT,JONATHAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 07/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: PINEDA

DATE/TIME OF TEST: 07/25/2018 01:35:05

Diagnostics Check	OK	01:23	
Air Blank	0.000	01:24	
Control Test	0.078	01:24	
Air Blank	0.000	01:24	
Subject Sample	# 1	01:28	No Sample Provided
Breath Volume	0.000	01:28	
Air Blank	0.000	01:28	
Air Blank	0.000	01:30	
Subject Sample	# 2	01:33	No Sample Provided
Breath Volume	0.000	01:33	
Air Blank	0.000	01:34	
Control Test	0.078	01:34	
Air Blank	0.000	01:34	
Diagnostics Check	OK	01:35	

Data Download Date/Time: 08/21/2018 12:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,CHAD,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: REED,CHAD
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 07/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: LINDSAY III

DATE/TIME OF TEST: 07/25/2018 22:03:07

Diagnostics Check	OK	21:56
Air Blank	0.000	21:57
Control Test	0.078	21:57
Air Blank	0.000	21:57
Subject Sample # 1	0.177	21:58
Breath Volume	2.523	21:58
Air Blank	0.000	21:59
Air Blank	0.000	22:00
Subject Sample # 2	0.173	22:01
Breath Volume	2.449	22:01
Air Blank	0.000	22:02
Control Test	0.076	22:02
Air Blank	0.000	22:03
Diagnostics Check	OK	22:03

Data Download Date/Time: 08/21/2018 12:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,CHAD,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: HURT,JONATHAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 07/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:31

SUBJECT'S LAST NAME: BIERNACKI

DATE/TIME OF TEST: 07/26/2018 21:07:17

Diagnostics Check	OK	21:00
Air Blank	0.000	21:00
Control Test	0.078	21:00
Air Blank	0.000	21:01
Subject Sample # 1	0.158	21:02
Breath Volume	3.523	21:02
Air Blank	0.000	21:03
Air Blank	0.000	21:05
Subject Sample # 2	0.150	21:05
Breath Volume	2.808	21:05
Air Blank	0.000	21:06
Control Test	0.078	21:06
Air Blank	0.000	21:07
Diagnostics Check	OK	21:07

Data Download Date/Time: 08/21/2018 12:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,CHAD,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: HURT,JONATHAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 07/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: CRONIN

DATE/TIME OF TEST: 07/27/2018 00:48:18

Diagnostics Check	OK	00:36	
Air Blank	0.000	00:37	
Control Test	0.077	00:37	
Air Blank	0.000	00:38	
Subject Sample	# 1	00:41	No Sample Provided
Breath Volume	0.000	00:41	
Air Blank	0.000	00:41	
Air Blank	0.000	00:43	
Subject Sample	# 2	00:46	No Sample Provided
Breath Volume	0.000	00:46	
Air Blank	0.000	00:47	
Control Test	0.078	00:47	
Air Blank	0.000	00:48	
Diagnostics Check	OK	00:48	

Data Download Date/Time: 08/21/2018 12:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,CHAD,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: REED,CHAD
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 07/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 07/27/2018 21:58:10

Diagnostics Check	OK	21:51
Air Blank	0.000	21:52
Control Test	0.078	21:52
Air Blank	0.000	21:52
Subject Sample # 1	0.000	21:53
Breath Volume	2.570	21:53
Air Blank	0.000	21:54
Air Blank	0.000	21:56
Subject Sample # 2	0.000	21:56
Breath Volume	2.132	21:56
Air Blank	0.000	21:57
Control Test	0.077	21:57
Air Blank	0.000	21:58
Diagnostics Check	OK	21:58

Data Download Date/Time: 08/21/2018 12:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTORO,CAMERON,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JETT,C
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 07/03/2018 21:17:09

Diagnostics Check	OK	21:10	
Air Blank	0.000	21:10	
Control Test	0.078	21:11	
Air Blank	0.000	21:11	
Subject Sample # 1	0.276	21:12	
Breath Volume	4.183	21:12	
Air Blank	0.000	21:12	
Air Blank	0.000	21:14	
Subject Sample # 2	0.298	21:15	
Breath Volume	6.546	21:15	
Air Blank		21:15	RFI Detect
Air Blank	0.000	21:16	
Control Test	0.079	21:16	
Air Blank	0.000	21:17	
Diagnostics Check	OK	21:17	

Data Download Date/Time: 07/26/2018 05:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTORO,CAMERON,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JETT,C
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 07/03/2018 21:27:01

Diagnostics Check	OK	21:20
Air Blank	0.000	21:21
Control Test	0.078	21:21
Air Blank	0.000	21:22
Subject Sample # 1	0.256	21:22
Breath Volume	3.234	21:22
Air Blank	0.000	21:23
Air Blank	0.000	21:25
Subject Sample # 2	0.267	21:25
Breath Volume	4.039	21:25
Air Blank	0.000	21:26
Control Test	0.079	21:26
Air Blank	0.000	21:26
Diagnostics Check	OK	21:27

Data Download Date/Time: 07/26/2018 05:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTORO,CAMERON,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: MILLS,D
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: HARRELL

DATE/TIME OF TEST: 07/12/2018 00:53:10

Diagnostics Check	OK	00:41	
Air Blank	0.000	00:42	
Control Test	0.079	00:42	
Air Blank	0.000	00:43	
Subject Sample	# 1	00:46	No Sample Provided
Breath Volume	0.000	00:46	
Air Blank	0.000	00:46	
Air Blank	0.000	00:48	
Subject Sample	# 2	00:51	No Sample Provided
Breath Volume	0.000	00:51	
Air Blank	0.000	00:52	
Control Test	0.079	00:52	
Air Blank	0.000	00:53	
Diagnostics Check	OK	00:53	

Data Download Date/Time: 07/26/2018 05:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARIDA,KORY,L
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: ALEXANDER,DAVID
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/10/2018 23:12:52

Diagnostics Check	OK	23:06
Air Blank	0.000	23:06
Control Test	0.079	23:07
Air Blank	0.000	23:07
Subject Sample # 1	0.000	23:08
Breath Volume	2.281	23:08
Air Blank	0.000	23:08
Air Blank	0.000	23:10
Subject Sample # 2	0.000	23:11
Breath Volume	2.023	23:11
Air Blank	0.000	23:11
Control Test	0.080	23:12
Air Blank	0.000	23:12
Diagnostics Check	OK	23:12

Data Download Date/Time: 07/26/2018 05:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARIDA,KORY,L
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: BELLEHUMEUR

DATE/TIME OF TEST: 07/12/2018 02:04:37

Diagnostics Check	OK	01:58
Air Blank	0.000	01:58
Control Test	0.079	01:59
Air Blank	0.000	01:59
Subject Sample # 1	0.250	02:00
Breath Volume	2.003	02:00
Air Blank	0.000	02:00
Air Blank	0.000	02:02
Subject Sample # 2	0.252	02:03
Breath Volume	2.566	02:03
Air Blank	0.000	02:03
Control Test	0.079	02:04
Air Blank	0.000	02:04
Diagnostics Check	OK	02:04

Data Download Date/Time: 07/26/2018 05:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTORO,C,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: PENA,A
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: MCKENZIE

DATE/TIME OF TEST: 07/14/2018 04:28:37

Diagnostics Check	OK	04:21
Air Blank	0.000	04:22
Control Test	0.079	04:22
Air Blank	0.000	04:23
Subject Sample # 1	0.150	04:24
Breath Volume	2.011	04:24
Air Blank	0.000	04:24
Air Blank	0.000	04:26
Subject Sample # 2	0.154	04:27
Breath Volume	1.285	04:27
Air Blank	0.000	04:27
Control Test	0.080	04:28
Air Blank	0.000	04:28
Diagnostics Check	OK	04:28

Data Download Date/Time: 07/26/2018 05:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROOKS,E,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: SMITH,J
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 13:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:27

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 07/15/2018 15:01:16

Diagnostics Check	OK	14:54
Air Blank	0.000	14:55
Control Test	0.079	14:55
Air Blank	0.000	14:55
Subject Sample # 1	0.097	14:56
Breath Volume	3.402	14:56
Air Blank	0.000	14:57
Air Blank	0.000	14:59
Subject Sample # 2	0.095	14:59
Breath Volume	4.484	14:59
Air Blank	0.000	15:00
Control Test	0.078	15:00
Air Blank	0.000	15:01
Diagnostics Check	OK	15:01

Data Download Date/Time: 07/26/2018 05:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENTZ,JARROD,C
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: BRICKEY,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:07

SUBJECT'S LAST NAME: GRILL

DATE/TIME OF TEST: 07/15/2018 19:39:32

Diagnostics Check	OK	19:33
Air Blank	0.000	19:33
Control Test	0.079	19:34
Air Blank	0.000	19:34
Subject Sample # 1	0.230	19:35
Breath Volume	1.312	19:35
Air Blank	0.000	19:35
Air Blank	0.000	19:37
Subject Sample # 2	0.238	19:38
Breath Volume	1.230	19:38
Air Blank	0.000	19:38
Control Test	0.078	19:38
Air Blank	0.000	19:39
Diagnostics Check	OK	19:39

Data Download Date/Time: 07/26/2018 05:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENTZ,JARROD,C
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: MCCLAIN,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: KELLEY

DATE/TIME OF TEST: 07/17/2018 05:32:16

Diagnostics Check	OK	05:26
Air Blank	0.000	05:26
Control Test	0.079	05:26
Air Blank	0.000	05:27
Subject Sample # 1	0.000	05:28
Breath Volume	1.445	05:28
Air Blank	0.000	05:28
Air Blank	0.000	05:30
Subject Sample # 2	0.000	05:30
Breath Volume	1.839	05:30
Air Blank	0.000	05:31
Control Test	0.079	05:31
Air Blank	0.000	05:32
Diagnostics Check	OK	05:32

Data Download Date/Time: 07/26/2018 05:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENTZ,JARROD,C
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: MILLER,K
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 07/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: PICKLO

DATE/TIME OF TEST: 07/19/2018 23:34:43

Diagnostics Check	OK	23:28
Air Blank	0.000	23:28
Control Test	0.078	23:29
Air Blank	0.000	23:29
Subject Sample # 1	0.171	23:30
Breath Volume	1.242	23:30
Air Blank	0.000	23:30
Air Blank	0.000	23:32
Subject Sample # 2	0.162	23:33
Breath Volume	1.921	23:33
Air Blank	0.000	23:33
Control Test	0.079	23:34
Air Blank	0.000	23:34
Diagnostics Check	OK	23:34

Data Download Date/Time: 10/16/2018 11:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARWOOD,CHARLES,M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: HARWOOD,CHARLES
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 07/17/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: ADMIN
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: SORRELLS

DATE/TIME OF TEST: 07/31/2018 18:30:23

Diagnostics Check	OK	18:29	
Air Blank	0.000	18:29	
Control Test	0.000	18:29	RFI Detect
Air Blank	0.000	18:30	

Data Download Date/Time: 10/16/2018 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARWOOD, CHARLES, M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: HARWOOD, CHARLES
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 07/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: ADMIN
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: SORRELLS

DATE/TIME OF TEST: 07/31/2018 18:38:43

Diagnostics Check	OK	18:32
Air Blank	0.000	18:33
Control Test	0.079	18:33
Air Blank	0.000	18:33
Subject Sample # 1	0.125	18:34
Breath Volume	2.785	18:34
Air Blank	0.000	18:34
Air Blank	0.000	18:36
Subject Sample # 2	0.112	18:37
Breath Volume	1.558	18:37
Air Blank	0.000	18:37
Control Test	0.079	18:38
Air Blank	0.000	18:38
Diagnostics Check	OK	18:38

Data Download Date/Time: 10/16/2018 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARWOOD, CHARLES, M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: HARWOOD, CHARLES
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 07/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: ADMIN
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: MEEK

DATE/TIME OF TEST: 07/31/2018 19:03:06

Diagnostics Check	OK	18:56
Air Blank	0.000	18:57
Control Test	0.079	18:57
Air Blank	0.000	18:58
Subject Sample # 1	0.063	18:58
Breath Volume	2.316	18:58
Air Blank	0.000	18:59
Air Blank	0.000	19:01
Subject Sample # 2	0.063	19:01
Breath Volume	2.808	19:01
Air Blank	0.000	19:02
Control Test	0.079	19:02
Air Blank	0.000	19:03
Diagnostics Check	OK	19:03

Data Download Date/Time: 10/16/2018 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARWOOD,CHARLES,M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: HARWOOD,CHARLES
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 07/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: ADMIN
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 07/31/2018 19:12:10

Diagnostics Check	OK	19:05
Air Blank	0.000	19:06
Control Test	0.078	19:06
Air Blank	0.000	19:06
Subject Sample # 1	0.000	19:07
Breath Volume	2.906	19:07
Air Blank	0.000	19:07
Air Blank	0.000	19:09
Subject Sample # 2	0.000	19:10
Breath Volume	3.027	19:10
Air Blank	0.000	19:11
Control Test	0.078	19:11
Air Blank	0.000	19:12
Diagnostics Check	OK	19:12

Data Download Date/Time: 10/16/2018 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARWOOD, CHARLES, M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: 2C09, HARWOOD
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 07/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: ADMIN
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: STONE

DATE/TIME OF TEST: 07/31/2018 19:23:08

Diagnostics Check	OK	19:17
Air Blank	0.000	19:17
Control Test	0.079	19:17
Air Blank	0.000	19:18
Subject Sample # 1	0.000	19:18
Breath Volume	3.027	19:18
Air Blank	0.000	19:19
Air Blank	0.000	19:21
Subject Sample # 2	0.000	19:21
Breath Volume	2.925	19:21
Air Blank	0.000	19:22
Control Test	0.078	19:22
Air Blank	0.000	19:23
Diagnostics Check	OK	19:23

Data Download Date/Time: 10/16/2018 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001073

INSTRUMENT REGISTERED TO: DADE CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEITH,HOOLEY,L
BREATH TEST OPERATOR AGENCY: Dade City PD
ARRESTING OFFICER: KEITH,HOOLEY
ARRESTING OFFICER AGENCY: Dade City PD
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A3
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: MEZA

DATE/TIME OF TEST: 07/21/2018 01:11:54

Diagnostics Check	OK	01:05
Air Blank	0.000	01:05
Control Test	0.078	01:05
Air Blank	0.000	01:06
Subject Sample # 1	0.074	01:07
Breath Volume	3.226	01:07
Air Blank	0.000	01:07
Air Blank	0.000	01:09
Subject Sample # 2	0.075	01:10
Breath Volume	2.863	01:10
Air Blank	0.000	01:11
Control Test	0.080	01:11
Air Blank	0.000	01:11
Diagnostics Check	OK	01:11

Data Download Date/Time: 08/16/2018 17:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAUGHERTY,DANIEL,M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: NIXON,BERRY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: WINN

DATE/TIME OF TEST: 07/04/2018 23:05:04

Diagnostics Check	OK	22:56
Air Blank	0.000	22:57
Control Test	0.078	22:57
Air Blank	0.000	22:58
Subject Sample # 1	0.000	23:00
Breath Volume	2.902	23:00
Air Blank	0.000	23:00
Air Blank	0.000	23:02
Subject Sample # 2	0.000	23:03
Breath Volume	2.222	23:03
Air Blank	0.000	23:04
Control Test	0.077	23:04
Air Blank	0.000	23:04
Diagnostics Check	OK	23:05

Data Download Date/Time: 07/11/2018 14:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,BARRY,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: NIXON,BARRY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: MILLS

DATE/TIME OF TEST: 07/15/2018 00:31:56

Diagnostics Check	OK	00:25
Air Blank	0.000	00:26
Control Test	0.078	00:26
Air Blank	0.000	00:27
Subject Sample # 1	0.087	00:27
Breath Volume	3.539	00:27
Air Blank	0.000	00:28
Air Blank	0.000	00:30
Subject Sample # 2	0.085	00:30
Breath Volume	2.136	00:30
Air Blank	0.000	00:31
Control Test	0.077	00:31
Air Blank	0.000	00:31
Diagnostics Check	OK	00:31

Data Download Date/Time: 08/03/2018 14:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,BARRY,M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: GRIFFIN,RAEFORD
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 07/20/2018 03:22:13

Diagnostics Check	OK	03:16
Air Blank	0.000	03:16
Control Test	0.079	03:16
Air Blank	0.000	03:17
Subject Sample # 1	0.272	03:17
Breath Volume	3.566	03:17
Air Blank	0.000	03:18
Air Blank	0.000	03:20
Subject Sample # 2	0.279	03:20
Breath Volume	3.550	03:20
Air Blank	0.000	03:21
Control Test	0.078	03:21
Air Blank	0.000	03:22
Diagnostics Check	OK	03:22

Data Download Date/Time: 08/03/2018 14:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CRUM,ZACKARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:54

SUBJECT'S LAST NAME: PEABODY

DATE/TIME OF TEST: 07/04/2018 22:18:22

Diagnostics Check	OK	22:15	
Air Blank	0.000	22:15	
Control Test	0.078	22:16	
Air Blank	0.000	22:16	
Subject Sample	# 1	22:16	Subject Test Refused
Breath Volume	0.000	22:16	
Air Blank	0.000	22:17	
Control Test	0.079	22:17	
Air Blank	0.000	22:18	
Diagnostics Check	OK	22:18	

Data Download Date/Time: 07/11/2018 10:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA,VLADIMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: MISSANA

DATE/TIME OF TEST: 07/05/2018 03:31:42

Diagnostics Check	OK	03:25
Air Blank	0.000	03:25
Control Test	0.079	03:26
Air Blank	0.000	03:26
Subject Sample # 1	0.242	03:27
Breath Volume	4.250	03:27
Air Blank	0.000	03:27
Air Blank	0.000	03:29
Subject Sample # 2	0.249	03:30
Breath Volume	3.781	03:30
Air Blank	0.000	03:30
Control Test	0.078	03:31
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time: 07/11/2018 10:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WALDO,AARON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: TOLBERT

DATE/TIME OF TEST: 07/07/2018 01:29:13

Diagnostics Check	OK	01:22
Air Blank	0.000	01:23
Control Test	0.080	01:23
Air Blank	0.000	01:24
Subject Sample # 1	0.178	01:24
Breath Volume	1.613	01:24
Air Blank	0.000	01:25
Air Blank	0.000	01:27
Subject Sample # 2	0.165	01:27
Breath Volume	1.214	01:27
Air Blank	0.000	01:28
Control Test	0.078	01:28
Air Blank	0.000	01:29
Diagnostics Check	OK	01:29

Data Download Date/Time: 07/11/2018 10:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDO,AARON,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KARAYIANES,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: HASSAN

DATE/TIME OF TEST: 07/07/2018 03:25:34

Diagnostics Check	OK	03:22	
Air Blank	0.000	03:23	
Control Test	0.079	03:23	
Air Blank	0.000	03:23	
Subject Sample	# 1	03:24	Subject Test Refused
Breath Volume	0.000	03:24	
Air Blank	0.000	03:24	
Control Test	0.078	03:24	
Air Blank	0.000	03:25	
Diagnostics Check	OK	03:25	

Data Download Date/Time: 07/11/2018 10:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CRUM,ZACKARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 07/07/2018 04:20:58

Diagnostics Check	OK	04:18	
Air Blank	0.000	04:18	
Control Test	0.078	04:18	
Air Blank	0.000	04:19	
Subject Sample	# 1	04:19	Subject Test Refused
Breath Volume	0.000	04:19	
Air Blank	0.000	04:20	
Control Test	0.079	04:20	
Air Blank	0.000	04:20	
Diagnostics Check	OK	04:20	

Data Download Date/Time: 07/11/2018 10:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDO,AARON,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KARAYIANES,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: ZARK

DATE/TIME OF TEST: 07/07/2018 22:34:56

Diagnostics Check	OK	22:27
Air Blank	0.000	22:28
Control Test	0.079	22:28
Air Blank	0.000	22:29
Subject Sample # 1	0.041	22:30
Breath Volume	1.222	22:30
Air Blank	0.000	22:30
Air Blank	0.000	22:32
Subject Sample # 2	0.043	22:33
Breath Volume	2.085	22:33
Air Blank	0.000	22:34
Control Test	0.079	22:34
Air Blank	0.000	22:34
Diagnostics Check	OK	22:34

Data Download Date/Time: 07/11/2018 10:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA,VLADIMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: BACHELER

DATE/TIME OF TEST: 07/13/2018 21:39:33

Diagnostics Check	OK	21:32
Air Blank	0.000	21:33
Control Test	0.079	21:33
Air Blank	0.000	21:34
Subject Sample # 1	0.000	21:34
Breath Volume	1.425	21:34
Air Blank	0.000	21:35
Air Blank	0.000	21:37
Subject Sample # 2	0.000	21:38
Breath Volume	3.113	21:38
Air Blank	0.000	21:38
Control Test	0.079	21:38
Air Blank	0.000	21:39
Diagnostics Check	OK	21:39

Data Download Date/Time: 08/21/2018 09:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUM,ZACKARY,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA,VLADAMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: OLIVER

DATE/TIME OF TEST: 07/14/2018 00:40:24

Diagnostics Check	OK	00:32
Air Blank	0.000	00:33
Control Test	0.079	00:33
Air Blank	0.000	00:34
Subject Sample # 1	0.127	00:35
Breath Volume	1.625	00:35
Air Blank	0.000	00:36
Air Blank	0.000	00:38
Subject Sample # 2	0.125	00:38
Breath Volume	1.613	00:38
Air Blank	0.000	00:39
Control Test	0.079	00:39
Air Blank	0.000	00:40
Diagnostics Check	OK	00:40

Data Download Date/Time: 08/21/2018 09:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KARAYIANES,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: DE BAKKER

DATE/TIME OF TEST: 07/14/2018 02:10:31

Diagnostics Check	OK	02:03
Air Blank	0.000	02:04
Control Test	0.078	02:04
Air Blank	0.000	02:04
Subject Sample # 1	0.222	02:05
Breath Volume	2.929	02:05
Air Blank	0.000	02:06
Air Blank	0.000	02:08
Subject Sample # 2	0.237	02:08
Breath Volume	3.390	02:08
Air Blank	0.000	02:09
Control Test	0.077	02:09
Air Blank	0.000	02:10
Diagnostics Check	OK	02:10

Data Download Date/Time: 08/21/2018 09:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEISKOPF,MICHAEL,D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA,VLADAMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: WARREN

DATE/TIME OF TEST: 07/14/2018 03:50:24

Diagnostics Check	OK	03:44
Air Blank	0.000	03:44
Control Test	0.078	03:44
Air Blank	0.000	03:45
Subject Sample # 1	0.173	03:46
Breath Volume	1.929	03:46
Air Blank	0.000	03:46
Air Blank	0.000	03:48
Subject Sample # 2	0.173	03:48
Breath Volume	1.640	03:48
Air Blank	0.000	03:49
Control Test	0.078	03:49
Air Blank	0.000	03:50
Diagnostics Check	OK	03:50

Data Download Date/Time: 08/21/2018 09:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CRUM,ZACKARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: STYLES

DATE/TIME OF TEST: 07/21/2018 03:41:38

Diagnostics Check	OK	03:38	
Air Blank	0.000	03:39	
Control Test	0.081	03:39	
Air Blank	0.000	03:40	
Subject Sample	# 1	03:40	Subject Test Refused
Breath Volume	0.000	03:40	
Air Blank	0.000	03:40	
Control Test	0.080	03:41	
Air Blank	0.000	03:41	
Diagnostics Check	OK	03:41	

Data Download Date/Time: 08/21/2018 09:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA,VLADIMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: PUGH

DATE/TIME OF TEST: 07/28/2018 02:12:27

Diagnostics Check	OK	02:03	
Air Blank	0.000	02:04	
Control Test	0.079	02:04	
Air Blank	0.000	02:05	
Subject Sample	# 1	02:08	No Sample Provided
Breath Volume	0.000	02:08	
Air Blank	0.000	02:08	
Air Blank	0.000	02:10	
Subject Sample	# 2	02:11	Subject Test Refused
Breath Volume	0.000	02:11	
Air Blank	0.000	02:11	
Control Test	0.079	02:11	
Air Blank	0.000	02:12	
Diagnostics Check	OK	02:12	

Data Download Date/Time: 08/21/2018 09:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CRUM,ZACKARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: GREGORY

DATE/TIME OF TEST: 07/29/2018 01:23:48

Diagnostics Check	OK	01:20	
Air Blank	0.000	01:21	
Control Test	0.079	01:21	
Air Blank	0.000	01:22	
Subject Sample	# 1	01:22	Subject Test Refused
Breath Volume	0.000	01:22	
Air Blank	0.000	01:22	
Control Test	0.079	01:23	
Air Blank	0.000	01:23	
Diagnostics Check	OK	01:23	

Data Download Date/Time: 08/21/2018 09:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASTINGS,KENNETH,S
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: HASTINGS,KENNETH
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: DALLAS

DATE/TIME OF TEST: 07/01/2018 00:06:42

Diagnostics Check	OK	00:03	
Air Blank	0.000	00:04	
Control Test	0.080	00:04	
Air Blank	0.000	00:05	
Subject Sample	# 1	00:05	Subject Test Refused
Breath Volume	0.000	00:05	
Air Blank	0.000	00:05	
Control Test	0.080	00:06	
Air Blank	0.000	00:06	
Diagnostics Check	OK	00:06	

Data Download Date/Time: 07/03/2018 13:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: SILVERSTEIN,CHRISTINA

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17817080A3

DRY GAS EXPIRY DATE: 08/05/2019

VIOLATION CODE: DUI

ARREST TIME: 02:04

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: HECKMAN

DATE/TIME OF TEST: 07/01/2018 02:53:51

Diagnostics Check	OK	02:46
Air Blank	0.000	02:46
Control Test	0.080	02:46
Air Blank	0.000	02:47
Subject Sample # 1	0.205	02:49
Breath Volume	2.308	02:49
Air Blank	0.000	02:49
Air Blank	0.000	02:51
Subject Sample # 2	0.212	02:52
Breath Volume	2.238	02:52
Air Blank	0.000	02:52
Control Test	0.079	02:53
Air Blank	0.000	02:53
Diagnostics Check	OK	02:53

Data Download Date/Time: 07/03/2018 13:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HASTINGS,KENNETH,S

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: HASTINGS,KENNETH

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 07/03/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17817080A3

DRY GAS EXPIRY DATE: 08/05/2019

VIOLATION CODE: DUI

ARREST TIME: 22:59

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: BOUTWELL

DATE/TIME OF TEST: 07/04/2018 23:48:58

Diagnostics Check	OK	23:45	
Air Blank	0.000	23:46	
Control Test	0.080	23:46	
Air Blank	0.000	23:47	
Subject Sample	# 1	23:47	Subject Test Refused
Breath Volume	0.000	23:47	
Air Blank	0.000	23:48	
Control Test	0.081	23:48	
Air Blank	0.000	23:48	
Diagnostics Check	OK	23:48	

Data Download Date/Time: 08/20/2018 10:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: SILVERSTEIN,CHRISTINA

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 07/03/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17817080A3

DRY GAS EXPIRY DATE: 08/05/2019

VIOLATION CODE: DUI

ARREST TIME: 03:36

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: EVOLA

DATE/TIME OF TEST: 07/07/2018 04:20:25

Diagnostics Check	OK	04:14
Air Blank	0.000	04:14
Control Test	0.081	04:14
Air Blank	0.000	04:15
Subject Sample # 1	0.102	04:15
Breath Volume	3.578	04:15
Air Blank	0.000	04:16
Air Blank	0.000	04:18
Subject Sample # 2	0.100	04:18
Breath Volume	3.824	04:18
Air Blank	0.000	04:19
Control Test	0.080	04:19
Air Blank	0.000	04:20
Diagnostics Check	OK	04:20

Data Download Date/Time: 08/20/2018 10:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,4L26
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 07/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: ROOT

DATE/TIME OF TEST: 07/08/2018 00:33:31

Diagnostics Check	OK	00:30	
Air Blank	0.000	00:31	
Control Test	0.080	00:31	
Air Blank	0.000	00:31	
Subject Sample	# 1	00:32	Subject Test Refused
Breath Volume	0.000	00:32	
Air Blank	0.000	00:32	
Control Test	0.080	00:32	
Air Blank	0.000	00:33	
Diagnostics Check	OK	00:33	

Data Download Date/Time: 08/20/2018 10:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HASTINGS,KENNETH,S

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: HASTINGS,KENNETH

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 07/03/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17817080A3

DRY GAS EXPIRY DATE: 08/05/2019

VIOLATION CODE: DUI

ARREST TIME: 03:44

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: WAINSCOTT

DATE/TIME OF TEST: 07/08/2018 04:31:54

Diagnostics Check	OK	04:25
Air Blank	0.000	04:25
Control Test	0.080	04:26
Air Blank	0.000	04:26
Subject Sample # 1	0.128	04:27
Breath Volume	1.769	04:27
Air Blank	0.000	04:27
Air Blank	0.000	04:29
Subject Sample # 2	0.139	04:30
Breath Volume	1.687	04:30
Air Blank	0.000	04:30
Control Test	0.080	04:31
Air Blank	0.000	04:31
Diagnostics Check	OK	04:31

Data Download Date/Time: 08/20/2018 10:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICE,JOHN,B
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: RICE,JOHN
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 07/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: ADMIN
ARREST TIME: 15:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:14

SUBJECT'S LAST NAME: GRASSIA

DATE/TIME OF TEST: 07/08/2018 17:41:45

Diagnostics Check	OK	17:34
Air Blank	0.000	17:35
Control Test	0.081	17:35
Air Blank	0.000	17:36
Subject Sample # 1	0.000	17:37
Breath Volume	2.601	17:37
Air Blank	0.000	17:37
Air Blank	0.000	17:39
Subject Sample # 2	0.000	17:40
Breath Volume	2.593	17:40
Air Blank	0.000	17:40
Control Test	0.080	17:41
Air Blank	0.000	17:41
Diagnostics Check	OK	17:41

Data Download Date/Time: 08/20/2018 10:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CITRANO, ANTHONY,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: HOLM, M.

ARRESTING OFFICER AGENCY: Indian Shores PD

LAST AGENCY INSPECTION: 07/03/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17817080A3

DRY GAS EXPIRY DATE: 08/05/2019

VIOLATION CODE: DUI

ARREST TIME: 04:19

GENDER: Male

OBSERVATION PERIOD BEGAN: 05:50

SUBJECT'S LAST NAME: LATVALA

DATE/TIME OF TEST: 07/11/2018 06:17:58

Diagnostics Check	OK	06:10
Air Blank	0.000	06:11
Control Test	0.079	06:11
Air Blank	0.000	06:12
Subject Sample # 1	0.092	06:13
Breath Volume	3.378	06:13
Air Blank	0.000	06:14
Air Blank	0.000	06:16
Subject Sample # 2	0.091	06:16
Breath Volume	2.371	06:16
Air Blank	0.000	06:17
Control Test	0.079	06:17
Air Blank	0.000	06:17
Diagnostics Check	OK	06:17

Data Download Date/Time: 08/20/2018 10:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BALDIE,BRYAN,A
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: KWIATKOWSKA,SYLWIA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 07/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:11

SUBJECT'S LAST NAME: USERY

DATE/TIME OF TEST: 07/12/2018 17:38:27

Diagnostics Check	OK	17:31
Air Blank	0.000	17:32
Control Test	0.080	17:32
Air Blank	0.000	17:32
Subject Sample # 1	0.127	17:33
Breath Volume	1.429	17:33
Air Blank	0.000	17:34
Air Blank	0.000	17:36
Subject Sample # 2	0.127	17:36
Breath Volume	1.843	17:36
Air Blank	0.000	17:37
Control Test	0.080	17:37
Air Blank	0.000	17:38
Diagnostics Check	OK	17:38

Data Download Date/Time: 08/20/2018 10:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HASTINGS,KENNETH,S

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: HASTINGS,KENNETH

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 07/03/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17817080A3

DRY GAS EXPIRY DATE: 08/05/2019

VIOLATION CODE: DUI

ARREST TIME: 03:30

GENDER: Female

OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 07/14/2018 04:10:52

Diagnostics Check	OK	04:03
Air Blank	0.000	04:04
Control Test	0.080	04:04
Air Blank	0.000	04:05
Subject Sample # 1	0.199	04:05
Breath Volume	1.609	04:05
Air Blank	0.000	04:06
Air Blank	0.000	04:08
Subject Sample # 2	0.211	04:09
Breath Volume	1.960	04:09
Air Blank	0.000	04:09
Control Test	0.080	04:10
Air Blank	0.000	04:10
Diagnostics Check	OK	04:10

Data Download Date/Time: 08/20/2018 10:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HASTINGS,KENNETH,S

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: HASTINGS,KENNETH

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 07/03/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17817080A3

DRY GAS EXPIRY DATE: 08/05/2019

VIOLATION CODE: DUI

ARREST TIME: 22:10

GENDER: Female

OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: LLOYD

DATE/TIME OF TEST: 07/14/2018 23:03:26

Diagnostics Check	OK	22:51	
Air Blank	0.000	22:52	
Control Test	0.081	22:52	
Air Blank	0.000	22:53	
Subject Sample	# 1 0.209	22:56	Volume Not Met
Breath Volume	0.015	22:56	
Air Blank	0.000	22:56	
Air Blank	0.000	22:58	
Subject Sample	# 2 0.225	23:01	Volume Not Met
Breath Volume	0.714	23:01	
Air Blank	0.000	23:02	
Control Test	0.079	23:02	
Air Blank	0.000	23:03	
Diagnostics Check	OK	23:03	

Data Download Date/Time: 08/20/2018 10:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HASTINGS,KENNETH,S

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: HASTINGS,KENNETH

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 07/03/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17817080A3

DRY GAS EXPIRY DATE: 08/05/2019

VIOLATION CODE: DUI

ARREST TIME: 22:10

GENDER: Female

OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: LLOYD

DATE/TIME OF TEST: 07/14/2018 23:17:03

Diagnostics Check	OK	23:10
Air Blank	0.000	23:11
Control Test	0.079	23:11
Air Blank	0.000	23:12
Subject Sample # 1	0.222	23:12
Breath Volume	1.714	23:12
Air Blank	0.000	23:13
Air Blank	0.000	23:14
Subject Sample # 2	0.232	23:15
Breath Volume	1.593	23:15
Air Blank	0.000	23:16
Control Test	0.080	23:16
Air Blank	0.000	23:16
Diagnostics Check	OK	23:17

Data Download Date/Time: 08/20/2018 10:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HASTINGS,KENNETH,S

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: HASTINGS,KENNETH

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 07/03/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17817080A3

DRY GAS EXPIRY DATE: 08/05/2019

VIOLATION CODE: DUI

ARREST TIME: 02:48

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: CRAIG

DATE/TIME OF TEST: 07/15/2018 03:33:39

Diagnostics Check	OK	03:27
Air Blank	0.000	03:27
Control Test	0.080	03:27
Air Blank	0.000	03:28
Subject Sample # 1	0.247	03:29
Breath Volume	2.265	03:29
Air Blank	0.000	03:29
Air Blank	0.000	03:31
Subject Sample # 2	0.233	03:32
Breath Volume	2.105	03:32
Air Blank	0.000	03:32
Control Test	0.080	03:33
Air Blank	0.000	03:33
Diagnostics Check	OK	03:33

Data Download Date/Time: 08/20/2018 10:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HASTINGS,KENNETH,S

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: HASTINGS,KENNETH

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 07/03/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17817080A3

DRY GAS EXPIRY DATE: 08/05/2019

VIOLATION CODE: DUI

ARREST TIME: 04:23

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:36

SUBJECT'S LAST NAME: RUSSELL

DATE/TIME OF TEST: 07/21/2018 05:10:48

Diagnostics Check	OK	04:56	
Air Blank	0.000	04:57	
Control Test	0.080	04:57	
Air Blank	0.000	04:58	
Subject Sample	# 1	05:01	No Sample Provided
Breath Volume	0.000	05:01	
Air Blank	0.000	05:01	
Air Blank	0.000	05:03	
Subject Sample	# 2	0.110 05:05	
Breath Volume	2.519	05:05	
Air Blank	0.000	05:05	
Air Blank	0.000	05:07	
Subject Sample	# 3	0.105 05:09	
Breath Volume	1.996	05:09	
Air Blank	0.000	05:09	
Control Test	0.080	05:10	
Air Blank	0.000	05:10	
Diagnostics Check	OK	05:10	

Data Download Date/Time: 08/20/2018 10:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,L
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 07/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: CULLEN

DATE/TIME OF TEST: 07/22/2018 05:43:22

Diagnostics Check	OK	05:31	
Air Blank	0.000	05:31	
Control Test	0.081	05:31	
Air Blank	0.000	05:32	
Subject Sample	# 1	05:35	No Sample Provided
Breath Volume	0.000	05:35	
Air Blank	0.000	05:36	
Air Blank	0.000	05:38	
Subject Sample	# 2	0.114 05:38	
Breath Volume	2.300	05:38	
Air Blank	0.000	05:39	
Air Blank	0.000	05:41	
Subject Sample	# 3	0.113 05:41	
Breath Volume	2.515	05:41	
Air Blank	0.000	05:42	
Control Test	0.080	05:42	
Air Blank	0.000	05:43	
Diagnostics Check	OK	05:43	

Data Download Date/Time: 08/20/2018 10:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWES,ERIC,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LAWES,ERIC
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 07/01/2018 00:46:28

Diagnostics Check	OK	00:39
Air Blank	0.000	00:40
Control Test	0.078	00:40
Air Blank	0.000	00:40
Subject Sample # 1	0.142	00:41
Breath Volume	3.304	00:41
Air Blank	0.000	00:42
Air Blank	0.000	00:44
Subject Sample # 2	0.142	00:44
Breath Volume	3.246	00:44
Air Blank	0.000	00:45
Control Test	0.077	00:45
Air Blank	0.000	00:46
Diagnostics Check	OK	00:46

Data Download Date/Time: 07/27/2018 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER, ADAM, H
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: POTTER, ADAM
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: TERVOL

DATE/TIME OF TEST: 07/01/2018 04:09:16

Diagnostics Check	OK	03:57	
Air Blank	0.000	03:58	
Control Test	0.077	03:58	
Air Blank	0.000	03:59	
Subject Sample	# 1	04:02	No Sample Provided
Breath Volume	0.000	04:02	
Air Blank	0.000	04:02	
Air Blank	0.000	04:04	
Subject Sample	# 2	04:07	No Sample Provided
Breath Volume	0.000	04:07	
Air Blank	0.000	04:08	
Control Test	0.077	04:08	
Air Blank	0.000	04:09	
Diagnostics Check	OK	04:09	

Data Download Date/Time: 07/27/2018 10:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COKE,CLEVELAND,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: CARTER,JACOB
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: BUI
ARREST TIME: 21:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: PRESTON

DATE/TIME OF TEST: 07/02/2018 00:47:51

Diagnostics Check	OK	00:44	
Air Blank	0.000	00:44	
Control Test	0.078	00:44	
Air Blank	0.000	00:45	
Subject Sample	# 1	00:46	Subject Test Refused
Breath Volume	0.000	00:46	
Air Blank	0.000	00:46	
Control Test	0.079	00:47	
Air Blank	0.000	00:47	
Diagnostics Check	OK	00:47	

Data Download Date/Time: 07/27/2018 10:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COKE,CLEVELAND,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: RESENDEZ,GILBRANDO
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: WASMUND

DATE/TIME OF TEST: 07/02/2018 00:54:33

Diagnostics Check	OK	00:53	
Air Blank	0.000	00:53	
		00:54	RFI Detect
Air Blank	0.000	00:54	

Data Download Date/Time: 07/27/2018 10:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COKE,CLEVELAND,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: RESENDEZ,GILBRANDO
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: WASMUND

DATE/TIME OF TEST: 07/02/2018 01:00:23

Diagnostics Check	OK	00:57	
Air Blank	0.000	00:58	
Control Test	0.078	00:58	
Air Blank	0.000	00:58	
Subject Sample	# 1	00:59	Subject Test Refused
Breath Volume	0.000	00:59	
Air Blank	0.000	00:59	
Control Test	0.078	00:59	
Air Blank	0.000	01:00	
Diagnostics Check	OK	01:00	

Data Download Date/Time: 07/27/2018 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGOUGH,ADRIN,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: CLIETT,JEREMY
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 12:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:53

SUBJECT'S LAST NAME: MEEKS

DATE/TIME OF TEST: 07/03/2018 14:30:09

Diagnostics Check	OK	14:24
Air Blank	0.000	14:24
Control Test	0.077	14:24
Air Blank	0.000	14:25
Subject Sample # 1	0.000	14:25
Breath Volume	2.542	14:25
Air Blank	0.000	14:26
Air Blank	0.000	14:28
Subject Sample # 2	0.000	14:28
Breath Volume	4.117	14:28
Air Blank	0.000	14:29
Control Test	0.075	14:29
Air Blank	0.000	14:30
Diagnostics Check	OK	14:30

Data Download Date/Time: 07/27/2018 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUCAS,NATHANIEL,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LUCAS,NATHANIEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 15:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:20

SUBJECT'S LAST NAME: STEELE

DATE/TIME OF TEST: 07/03/2018 16:54:00

Diagnostics Check	OK	16:45
Air Blank	0.000	16:45
Control Test	0.077	16:46
Air Blank	0.000	16:46
Subject Sample # 1	0.137	16:49
Breath Volume	1.644	16:49
Air Blank	0.000	16:49
Air Blank	0.000	16:51
Subject Sample # 2	0.141	16:52
Breath Volume	1.960	16:52
Air Blank	0.000	16:53
Control Test	0.075	16:53
Air Blank	0.000	16:53
Diagnostics Check	OK	16:54

Data Download Date/Time: 07/27/2018 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,RANDY,O
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WILSON,RANDY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: CELENZA

DATE/TIME OF TEST: 07/04/2018 04:53:56

Diagnostics Check	OK	04:42	
Air Blank	0.000	04:42	
Control Test	0.078	04:43	
Air Blank	0.000	04:43	
Subject Sample	# 1	04:46	No Sample Provided
Breath Volume	0.000	04:46	
Air Blank	0.000	04:47	
Air Blank	0.000	04:49	
Subject Sample	# 2	04:52	No Sample Provided
Breath Volume	0.000	04:52	
Air Blank	0.000	04:53	
Control Test	0.078	04:53	
Air Blank	0.000	04:53	
Diagnostics Check	OK	04:53	

Data Download Date/Time: 07/27/2018 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER, ADAM, H
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: EARLS, CHRISTIAN
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:56

SUBJECT'S LAST NAME: SURRENCY

DATE/TIME OF TEST: 07/04/2018 20:37:38

Diagnostics Check	OK	20:31
Air Blank	0.000	20:31
Control Test	0.078	20:32
Air Blank	0.000	20:32
Subject Sample # 1	0.130	20:33
Breath Volume	2.429	20:33
Air Blank	0.000	20:33
Air Blank	0.000	20:35
Subject Sample # 2	0.128	20:36
Breath Volume	2.136	20:36
Air Blank	0.000	20:36
Control Test	0.078	20:37
Air Blank	0.000	20:37
Diagnostics Check	OK	20:37

Data Download Date/Time: 07/27/2018 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWES,ERIC,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LAWES,ERIC
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: ARAUJO

DATE/TIME OF TEST: 07/04/2018 22:30:28

Diagnostics Check	OK	22:23
Air Blank	0.000	22:23
Control Test	0.078	22:24
Air Blank	0.000	22:24
Subject Sample # 1	0.081	22:25
Breath Volume	1.343	22:25
Air Blank	0.000	22:26
Air Blank	0.000	22:28
Subject Sample # 2	0.071	22:28
Breath Volume	1.562	22:28
Air Blank	0.000	22:29
Control Test	0.078	22:29
Air Blank	0.000	22:30
Diagnostics Check	OK	22:30

Data Download Date/Time: 07/27/2018 10:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWES,ERIC,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ROBERTS,DALE
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: KLOSTERMAN

DATE/TIME OF TEST: 07/04/2018 23:18:49

Diagnostics Check	OK	23:07
Air Blank	0.000	23:07
Control Test	0.078	23:08
Air Blank	0.000	23:08
Subject Sample # 1	0.215	23:10
Breath Volume	2.187	23:10
Air Blank	0.000	23:11
Air Blank	0.000	23:12
Subject Sample # 2	0.189	23:13
Breath Volume	1.250	23:13
Air Blank	0.000	23:14
Air Blank	0.000	23:16
Subject Sample # 3	0.219	23:17
Breath Volume	2.449	23:17
Air Blank	0.000	23:17
Control Test	0.078	23:18
Air Blank	0.000	23:18
Diagnostics Check	OK	23:18

Data Download Date/Time: 07/27/2018 10:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWES,ERIC,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: CARTER,JACOB
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: ENGLE

DATE/TIME OF TEST: 07/05/2018 00:55:06

Diagnostics Check	OK	00:52	
Air Blank	0.000	00:52	
Control Test	0.077	00:52	
Air Blank	0.000	00:53	
Subject Sample	# 1	00:53	Subject Test Refused
Breath Volume	0.000	00:53	
Air Blank	0.000	00:54	
Control Test	0.078	00:54	
Air Blank	0.000	00:55	
Diagnostics Check	OK	00:55	

Data Download Date/Time: 07/27/2018 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIOSO, HELENA, D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: PRESENDIEU, RENAND
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 06:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:10

SUBJECT'S LAST NAME: NIOSO

DATE/TIME OF TEST: 07/05/2018 08:42:49

Diagnostics Check	OK	08:31	
Air Blank	0.000	08:31	
Control Test	0.078	08:32	
Air Blank	0.000	08:32	
Subject Sample	# 1	08:35	No Sample Provided
Breath Volume	0.000	08:35	
Air Blank	0.000	08:36	
Air Blank	0.000	08:38	
Subject Sample	# 2	08:41	No Sample Provided
Breath Volume	0.000	08:41	
Air Blank	0.000	08:41	
Control Test	0.078	08:42	
Air Blank	0.000	08:42	
Diagnostics Check	OK	08:42	

Data Download Date/Time: 07/27/2018 10:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIOSO, HELENA, D
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: PRESENDIEU, RENAND
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 06:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:45

SUBJECT'S LAST NAME: N200324958270

DATE/TIME OF TEST: 07/05/2018 09:13:34

Diagnostics Check	OK	09:07
Air Blank	0.000	09:07
Control Test	0.078	09:08
Air Blank	0.000	09:08
Subject Sample # 1	0.070	09:09
Breath Volume	2.546	09:09
Air Blank	0.000	09:09
Air Blank	0.000	09:11
Subject Sample # 2	0.067	09:11
Breath Volume	2.476	09:11
Air Blank	0.000	09:12
Control Test	0.077	09:12
Air Blank	0.000	09:13
Diagnostics Check	OK	09:13

Data Download Date/Time: 07/27/2018 10:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCH,JOHN,M
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: CHURCH,JOHN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 15:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:23

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/05/2018 16:52:03

Diagnostics Check	OK	16:49	
Air Blank	0.000	16:49	
Control Test	0.078	16:49	
Air Blank	0.000	16:50	
Subject Sample	# 1	16:50	Subject Test Refused
Breath Volume	0.000	16:50	
Air Blank	0.000	16:51	
Control Test	0.078	16:51	
Air Blank	0.000	16:51	
Diagnostics Check	OK	16:52	

Data Download Date/Time: 07/27/2018 10:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWES,ERIC,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ROBERTS,DALE
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: SALVATORE

DATE/TIME OF TEST: 07/06/2018 00:14:46

Diagnostics Check	OK	00:05
Air Blank	0.000	00:05
Control Test	0.078	00:06
Air Blank	0.000	00:06
Subject Sample # 1	0.247	00:07
Breath Volume	2.363	00:07
Air Blank	0.000	00:08
Air Blank	0.000	00:10
Subject Sample # 2	0.243	00:13
Breath Volume	3.265	00:13
Air Blank	0.000	00:13
Control Test	0.078	00:14
Air Blank	0.000	00:14
Diagnostics Check	OK	00:14

Data Download Date/Time: 07/27/2018 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWES,ERIC,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WHITE,AMBER
ARRESTING OFFICER AGENCY: Bartow PD
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: BREWER

DATE/TIME OF TEST: 07/06/2018 04:36:06

Diagnostics Check	OK	04:29
Air Blank	0.000	04:29
Control Test	0.078	04:30
Air Blank	0.000	04:30
Subject Sample # 1	0.177	04:31
Breath Volume	1.671	04:31
Air Blank	0.000	04:31
Air Blank	0.000	04:33
Subject Sample # 2	0.180	04:34
Breath Volume	2.214	04:34
Air Blank	0.000	04:35
Control Test	0.077	04:35
Air Blank	0.000	04:36
Diagnostics Check	OK	04:36

Data Download Date/Time: 07/27/2018 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOFFORD,CLINT,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WOFFORD,CLINT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: HART

DATE/TIME OF TEST: 07/07/2018 01:10:18

Diagnostics Check	OK	01:03
Air Blank	0.000	01:03
Control Test	0.078	01:04
Air Blank	0.000	01:04
Subject Sample # 1	0.180	01:05
Breath Volume	2.007	01:05
Air Blank	0.000	01:05
Air Blank	0.000	01:07
Subject Sample # 2	0.174	01:08
Breath Volume	1.371	01:08
Air Blank	0.000	01:09
Control Test	0.077	01:09
Air Blank	0.000	01:10
Diagnostics Check	OK	01:10

Data Download Date/Time: 07/27/2018 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES, STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SYKES, STEVEN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:32

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 07/07/2018 19:02:32

Diagnostics Check	OK	18:56
Air Blank	0.000	18:56
Control Test	0.079	18:57
Air Blank	0.000	18:57
Subject Sample # 1	0.000	18:58
Breath Volume	1.218	18:58
Air Blank	0.000	18:58
Air Blank	0.000	19:00
Subject Sample # 2	0.000	19:01
Breath Volume	1.964	19:01
Air Blank	0.000	19:01
Control Test	0.078	19:01
Air Blank	0.000	19:02
Diagnostics Check	OK	19:02

Data Download Date/Time: 07/27/2018 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHRISTELLO,CLIFFORD,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: CHRISTELLO,CLIFFORD
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: BLANKENSHIP

DATE/TIME OF TEST: 07/07/2018 21:25:41

Diagnostics Check	OK	21:16	
Air Blank	0.000	21:16	
Control Test	0.077	21:17	
Air Blank	0.000	21:17	
Subject Sample	# 1 0.197	21:19	
Breath Volume	1.199	21:19	
Air Blank	0.000	21:20	
Air Blank	0.000	21:22	
Subject Sample	# 2	21:23	RFI Detect
Breath Volume	0.863	21:23	
Air Blank	0.000	21:24	
Control Test	0.076	21:25	
Air Blank	0.000	21:25	
Diagnostics Check	OK	21:25	

Data Download Date/Time: 07/27/2018 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHRISTELLO,CLIFFORD,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: CHRISTELLO,CLIFFORD
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: BLANKENSHIP

DATE/TIME OF TEST: 07/07/2018 21:36:46

Diagnostics Check	OK	21:28
Air Blank	0.000	21:29
Control Test	0.076	21:29
Air Blank	0.000	21:30
Subject Sample # 1	0.182	21:30
Breath Volume	2.074	21:30
Air Blank	0.000	21:31
Air Blank	0.000	21:32
Subject Sample # 2	0.186	21:35
Breath Volume	2.140	21:35
Air Blank	0.000	21:35
Control Test	0.077	21:36
Air Blank	0.000	21:36
Diagnostics Check	OK	21:36

Data Download Date/Time: 07/27/2018 10:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUCAS,NATHANIEL,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LUCAS,NATHANIEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: LESCHAK

DATE/TIME OF TEST: 07/07/2018 22:42:40

Diagnostics Check	OK	22:39	
Air Blank	0.000	22:40	
Control Test	0.078	22:40	
Air Blank	0.000	22:41	
Subject Sample	# 1	22:41	Subject Test Refused
Breath Volume	0.000	22:41	
Air Blank	0.000	22:41	
Control Test	0.078	22:42	
Air Blank	0.000	22:42	
Diagnostics Check	OK	22:42	

Data Download Date/Time: 07/27/2018 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES, STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SYKES, STEVEN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 07/08/2018 00:18:53

Diagnostics Check	OK	00:12
Air Blank	0.000	00:13
Control Test	0.078	00:13
Air Blank	0.000	00:14
Subject Sample # 1	0.164	00:14
Breath Volume	1.976	00:14
Air Blank	0.000	00:15
Air Blank	0.000	00:16
Subject Sample # 2	0.156	00:17
Breath Volume	1.562	00:17
Air Blank	0.000	00:17
Control Test	0.078	00:18
Air Blank	0.000	00:18
Diagnostics Check	OK	00:18

Data Download Date/Time: 07/27/2018 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOFFORD,CLINT,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WOFFORD,CLINT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:11

SUBJECT'S LAST NAME: HERNANDEZ-JIMENEZ

DATE/TIME OF TEST: 07/08/2018 05:41:07

Diagnostics Check	OK	05:34
Air Blank	0.000	05:34
Control Test	0.079	05:35
Air Blank	0.000	05:35
Subject Sample # 1	0.126	05:36
Breath Volume	2.660	05:36
Air Blank	0.000	05:37
Air Blank	0.000	05:38
Subject Sample # 2	0.121	05:39
Breath Volume	2.492	05:39
Air Blank	0.000	05:40
Control Test	0.079	05:40
Air Blank	0.000	05:41
Diagnostics Check	OK	05:41

Data Download Date/Time: 07/27/2018 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER, ADAM, H
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: POTTER, ADAM
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: LEACH

DATE/TIME OF TEST: 07/10/2018 01:30:26

Diagnostics Check	OK	01:22
Air Blank	0.000	01:23
Control Test	0.079	01:23
Air Blank	0.000	01:24
Subject Sample # 1	0.170	01:24
Breath Volume	1.882	01:24
Air Blank	0.000	01:25
Air Blank	0.000	01:27
Subject Sample # 2	0.152	01:28
Breath Volume	1.570	01:28
Air Blank	0.000	01:29
Control Test	0.078	01:29
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

Data Download Date/Time: 07/27/2018 10:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOFFORD,CLINT,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WOFFORD,CLINT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 07/13/2018 01:32:17

Diagnostics Check	OK	01:29	
Air Blank	0.000	01:29	
Control Test	0.078	01:30	
Air Blank	0.000	01:30	
Subject Sample	# 1	01:30	Subject Test Refused
Breath Volume	0.000	01:30	
Air Blank	0.000	01:31	
Control Test	0.079	01:31	
Air Blank	0.000	01:32	
Diagnostics Check	OK	01:32	

Data Download Date/Time: 07/27/2018 10:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWES,ERIC,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LAWES,ERIC
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 07/15/2018 02:41:16

Diagnostics Check	OK	02:40	
Air Blank	0.000	02:40	
Control Test	0.077	02:40	RFI Detect
Air Blank	0.000	02:41	

Data Download Date/Time: 07/27/2018 10:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWES,ERIC,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LAWES,ERIC
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 07/15/2018 02:54:14

Diagnostics Check	OK	02:42	
Air Blank	0.000	02:43	
Control Test	0.078	02:43	
Air Blank	0.000	02:44	
Subject Sample	# 1	02:47	No Sample Provided
Breath Volume	0.000	02:47	
Air Blank	0.000	02:47	
Air Blank	0.000	02:49	
Subject Sample	# 2	02:52	No Sample Provided
Breath Volume	0.000	02:52	
Air Blank	0.000	02:53	
Control Test	0.078	02:53	
Air Blank	0.000	02:54	
Diagnostics Check	OK	02:54	

Data Download Date/Time: 07/27/2018 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER, ADAM, H
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: POTTER, ADAM
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: SILVIA

DATE/TIME OF TEST: 07/15/2018 03:08:35

Diagnostics Check	OK	03:00
Air Blank	0.000	03:00
Control Test	0.079	03:00
Air Blank	0.000	03:01
Subject Sample # 1	0.171	03:01
Breath Volume	1.179	03:01
Air Blank	0.000	03:02
Air Blank	0.000	03:04
Subject Sample # 2	0.168	03:06
Breath Volume	1.160	03:06
Air Blank	0.000	03:07
Control Test	0.077	03:07
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

Data Download Date/Time: 07/27/2018 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWES,ERIC,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LAWES,ERIC
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 06:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:45

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 07/15/2018 07:17:35

Diagnostics Check	OK	07:10
Air Blank	0.000	07:11
Control Test	0.078	07:11
Air Blank	0.000	07:11
Subject Sample # 1	0.175	07:12
Breath Volume	2.175	07:12
Air Blank	0.000	07:13
Air Blank	0.000	07:15
Subject Sample # 2	0.166	07:16
Breath Volume	2.019	07:16
Air Blank	0.000	07:16
Control Test	0.077	07:16
Air Blank	0.000	07:17
Diagnostics Check	OK	07:17

Data Download Date/Time: 07/27/2018 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRESENDIEU, RENAND,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: PRESENDIEU, RENAND
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 08:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: JUAREZ

DATE/TIME OF TEST: 07/15/2018 22:20:20

Diagnostics Check	OK	22:13
Air Blank	0.000	22:14
Control Test	0.078	22:14
Air Blank	0.000	22:15
Subject Sample # 1	0.105	22:15
Breath Volume	1.289	22:15
Air Blank	0.000	22:16
Air Blank	0.000	22:18
Subject Sample # 2	0.107	22:18
Breath Volume	1.566	22:18
Air Blank	0.000	22:19
Control Test	0.077	22:19
Air Blank	0.000	22:20
Diagnostics Check	OK	22:20

Data Download Date/Time: 07/27/2018 10:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY, STEVEN, S
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: CARTER, JACOB
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 10:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: FARLEY

DATE/TIME OF TEST: 07/16/2018 00:55:16

Diagnostics Check	OK	00:52	
Air Blank	0.000	00:52	
Control Test	0.079	00:53	
Air Blank	0.000	00:53	
Subject Sample	# 1	00:53	Subject Test Refused
Breath Volume	0.000	00:53	
Air Blank	0.000	00:54	
Control Test	0.079	00:54	
Air Blank	0.000	00:55	
Diagnostics Check	OK	00:55	

Data Download Date/Time: 07/27/2018 10:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRINCE, CARMON, W
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: CRUZ, EDGARDO
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: PRINCE

DATE/TIME OF TEST: 07/16/2018 01:09:08

Diagnostics Check	OK	01:06	
Air Blank	0.000	01:06	
Control Test	0.078	01:07	
Air Blank	0.000	01:07	
Subject Sample	# 1	01:07	Subject Test Refused
Breath Volume	0.000	01:07	
Air Blank	0.000	01:08	
Control Test	0.078	01:08	
Air Blank	0.000	01:09	
Diagnostics Check	OK	01:09	

Data Download Date/Time: 07/27/2018 10:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELTON,XAVIER,V
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: CRUZ,EDUARDO
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: PRINCE

DATE/TIME OF TEST: 07/16/2018 01:32:23

Diagnostics Check	OK	01:29	
Air Blank	0.000	01:29	
Control Test	0.079	01:29	
Air Blank	0.000	01:30	
Subject Sample	# 1	01:31	Subject Test Refused
Breath Volume	0.000	01:31	
Air Blank	0.000	01:31	
Control Test	0.079	01:31	
Air Blank	0.000	01:32	
Diagnostics Check	OK	01:32	

Data Download Date/Time: 07/27/2018 10:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWES,ERIC,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LAWES,ERIC
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: JEFFRIES

DATE/TIME OF TEST: 07/16/2018 05:19:22

Diagnostics Check	OK	05:13
Air Blank	0.000	05:13
Control Test	0.077	05:13
Air Blank	0.000	05:14
Subject Sample # 1	0.000	05:15
Breath Volume	2.550	05:15
Air Blank	0.000	05:15
Air Blank	0.000	05:17
Subject Sample # 2	0.000	05:18
Breath Volume	2.707	05:18
Air Blank	0.000	05:18
Control Test	0.077	05:18
Air Blank	0.000	05:19
Diagnostics Check	OK	05:19

Data Download Date/Time: 07/27/2018 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUCAS,NATHANIEL,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LUCAS,NATHANIEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: VANESS

DATE/TIME OF TEST: 07/16/2018 18:36:12

Diagnostics Check	OK	18:29
Air Blank	0.000	18:30
Control Test	0.079	18:30
Air Blank	0.000	18:31
Subject Sample # 1	0.239	18:31
Breath Volume	2.257	18:31
Air Blank	0.000	18:32
Air Blank	0.000	18:34
Subject Sample # 2	0.243	18:34
Breath Volume	2.171	18:34
Air Blank	0.000	18:35
Control Test	0.077	18:35
Air Blank	0.000	18:36
Diagnostics Check	OK	18:36

Data Download Date/Time: 07/27/2018 10:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILLIPS,JORDAN,W
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: CLAROS,EDGAR
ARRESTING OFFICER AGENCY: Lake Wales PD
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: GABRIEL SALVADOR

DATE/TIME OF TEST: 07/17/2018 00:20:04

Diagnostics Check	OK	00:17	
Air Blank	0.000	00:17	
Control Test	0.078	00:18	
Air Blank	0.000	00:18	
Subject Sample	# 1	00:18	Subject Test Refused
Breath Volume	0.000	00:18	
Air Blank	0.000	00:19	
Control Test	0.078	00:19	
Air Blank	0.000	00:19	
Diagnostics Check	OK	00:20	

Data Download Date/Time: 07/27/2018 10:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELTON,XAVIER,V
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JONES,MICHEAL
ARRESTING OFFICER AGENCY: Lake Wales PD
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: KIDD

DATE/TIME OF TEST: 07/19/2018 04:07:41

Diagnostics Check	OK	04:01
Air Blank	0.000	04:01
Control Test	0.078	04:01
Air Blank	0.000	04:02
Subject Sample # 1	0.221	04:02
Breath Volume	3.445	04:02
Air Blank	0.000	04:03
Air Blank	0.000	04:05
Subject Sample # 2	0.217	04:05
Breath Volume	4.015	04:05
Air Blank	0.000	04:06
Control Test	0.075	04:07
Air Blank	0.000	04:07
Diagnostics Check	OK	04:07

Data Download Date/Time: 07/27/2018 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILLIPS,JORDAN,W
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: CARTER,JACOB
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: BUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: TOOLE

DATE/TIME OF TEST: 07/21/2018 01:01:41

Diagnostics Check	OK	00:54
Air Blank	0.000	00:55
Control Test	0.078	00:55
Air Blank	0.000	00:55
Subject Sample # 1	0.109	00:56
Breath Volume	1.367	00:56
Air Blank	0.000	00:57
Air Blank	0.000	00:59
Subject Sample # 2	0.120	01:00
Breath Volume	1.367	01:00
Air Blank	0.000	01:00
Control Test	0.076	01:01
Air Blank	0.000	01:01
Diagnostics Check	OK	01:01

Data Download Date/Time: 07/27/2018 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARREN,BILLY,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WARREN,BILLY
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: WARD

DATE/TIME OF TEST: 07/21/2018 04:27:08

Diagnostics Check	OK	04:20
Air Blank	0.000	04:21
Control Test	0.078	04:21
Air Blank	0.000	04:21
Subject Sample # 1	0.154	04:22
Breath Volume	4.210	04:22
Air Blank	0.000	04:23
Air Blank	0.000	04:24
Subject Sample # 2	0.149	04:25
Breath Volume	4.824	04:25
Air Blank	0.000	04:26
Control Test	0.077	04:26
Air Blank	0.000	04:27
Diagnostics Check	OK	04:27

Data Download Date/Time: 07/27/2018 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOFFORD,CLINT,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WOFFORD,CLINT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: APARICIO

DATE/TIME OF TEST: 07/21/2018 21:47:56

Diagnostics Check	OK	21:38
Air Blank	0.000	21:39
Control Test	0.078	21:39
Air Blank	0.000	21:39
Subject Sample # 1	0.148	21:41
Breath Volume	2.085	21:41
Air Blank	0.000	21:41
Air Blank	0.000	21:43
Subject Sample # 2	0.132	21:46
Breath Volume	1.609	21:46
Air Blank	0.000	21:47
Control Test	0.077	21:47
Air Blank	0.000	21:47
Diagnostics Check	OK	21:47

Data Download Date/Time: 07/27/2018 10:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWES,ERIC,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: PAUL,CORY
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:10

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/24/2018 06:58:23

Diagnostics Check	OK	06:51
Air Blank	0.000	06:52
Control Test	0.077	06:52
Air Blank	0.000	06:52
Subject Sample # 1	0.222	06:53
Breath Volume	3.820	06:53
Air Blank	0.000	06:54
Air Blank	0.000	06:56
Subject Sample # 2	0.219	06:56
Breath Volume	3.640	06:56
Air Blank	0.000	06:57
Control Test	0.076	06:57
Air Blank	0.000	06:58
Diagnostics Check	OK	06:58

Data Download Date/Time: 07/27/2018 10:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOFFORD,CLINT,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WOFFORD,CLINT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: CHRISTENSEN

DATE/TIME OF TEST: 07/27/2018 02:46:51

Diagnostics Check	OK	02:43	
Air Blank	0.000	02:43	
Control Test	0.078	02:43	
Air Blank	0.000	02:44	
Subject Sample	# 1	02:45	Interferent Detect
Breath Volume	2.234	02:45	
Air Blank	0.000	02:45	
Control Test	0.078	02:46	
Air Blank	0.000	02:46	
Diagnostics Check	OK	02:46	

Data Download Date/Time: 07/27/2018 10:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOFFORD,CLINT,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WOFFORD,CLINT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: CHRISTENSEN

DATE/TIME OF TEST: 07/27/2018 02:53:06

Diagnostics Check	OK	02:49	
Air Blank	0.000	02:50	
Control Test	0.077	02:50	
Air Blank	0.000	02:50	
Subject Sample	# 1	02:51	Interferent Detect
Breath Volume	2.105	02:51	
Air Blank	0.000	02:52	
Control Test	0.077	02:52	
Air Blank	0.000	02:53	
Diagnostics Check	OK	02:53	

Data Download Date/Time: 07/27/2018 10:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARQUHARSON,KEVIN,D
BREATH TEST OPERATOR AGENCY: Plantation PD
ARRESTING OFFICER: BEAUREGARD,JEFF
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 06/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A5
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: BARIENTO

DATE/TIME OF TEST: 07/05/2018 01:52:59

Diagnostics Check	OK	01:49	
Air Blank	0.000	01:50	
Control Test	0.079	01:50	
Air Blank	0.000	01:51	
Subject Sample	# 1	01:51	Subject Test Refused
Breath Volume	0.000	01:51	
Air Blank	0.000	01:52	
Control Test	0.079	01:52	
Air Blank	0.000	01:52	
Diagnostics Check	OK	01:52	

Data Download Date/Time: 07/26/2018 15:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARQUHARSON,KEVIN,D
BREATH TEST OPERATOR AGENCY: Plantation PD
ARRESTING OFFICER: FARQUHARSON,KEVIN
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 07/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A5
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: MARTINEZ GUILLEN

DATE/TIME OF TEST: 07/28/2018 06:36:15

Diagnostics Check	OK	06:33	
Air Blank	0.000	06:33	
Control Test	0.079	06:34	
Air Blank	0.000	06:34	
Subject Sample	# 1	06:34	Subject Test Refused
Breath Volume	0.000	06:34	
Air Blank	0.000	06:35	
Control Test	0.079	06:35	
Air Blank	0.000	06:36	
Diagnostics Check	OK	06:36	

Data Download Date/Time: 08/29/2018 20:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BROWN,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: BARREDA LANDA

DATE/TIME OF TEST: 07/20/2018 00:56:10

Diagnostics Check	OK	00:49
Air Blank	0.000	00:50
Control Test	0.079	00:50
Air Blank	0.000	00:50
Subject Sample # 1	0.220	00:51
Breath Volume	2.089	00:51
Air Blank	0.000	00:52
Air Blank	0.000	00:54
Subject Sample # 2	0.219	00:54
Breath Volume	2.136	00:54
Air Blank	0.000	00:55
Control Test	0.080	00:55
Air Blank	0.000	00:56
Diagnostics Check	OK	00:56

Data Download Date/Time: 08/17/2018 11:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: VAN CLEVE-MILLER

DATE/TIME OF TEST: 07/20/2018 03:45:56

Diagnostics Check	OK	03:38
Air Blank	0.000	03:39
Control Test	0.080	03:39
Air Blank	0.000	03:40
Subject Sample # 1	0.000	03:41
Breath Volume	1.835	03:41
Air Blank	0.000	03:42
Air Blank	0.000	03:44
Subject Sample # 2	0.000	03:44
Breath Volume	1.433	03:44
Air Blank	0.000	03:45
Control Test	0.079	03:45
Air Blank	0.000	03:45
Diagnostics Check	OK	03:45

Data Download Date/Time: 08/17/2018 11:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HARD,W
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: GREENE

DATE/TIME OF TEST: 07/20/2018 04:42:10

Diagnostics Check	OK	04:39	
Air Blank	0.000	04:39	
Control Test	0.080	04:40	
Air Blank	0.000	04:40	
Subject Sample	# 1	04:40	Subject Test Refused
Breath Volume	0.000	04:40	
Air Blank	0.000	04:41	
Control Test	0.080	04:41	
Air Blank	0.000	04:42	
Diagnostics Check	OK	04:42	

Data Download Date/Time: 08/17/2018 11:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: HAMWEY

DATE/TIME OF TEST: 07/20/2018 23:33:34

Diagnostics Check	OK	23:30	
Air Blank	0.000	23:31	
Control Test	0.079	23:31	
Air Blank	0.000	23:32	
Subject Sample	# 1	23:32	Subject Test Refused
Breath Volume	0.000	23:32	
Air Blank	0.000	23:32	
Control Test	0.079	23:32	
Air Blank	0.000	23:33	
Diagnostics Check	OK	23:33	

Data Download Date/Time: 08/17/2018 11:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BROWN,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: OLIVER

DATE/TIME OF TEST: 07/21/2018 01:20:48

Diagnostics Check	OK	01:18	
Air Blank	0.000	01:18	
Control Test	0.080	01:18	
Air Blank	0.000	01:19	
Subject Sample	# 1	01:19	Subject Test Refused
Breath Volume	0.000	01:19	
Air Blank	0.000	01:19	
Control Test	0.079	01:20	
Air Blank	0.000	01:20	
Diagnostics Check	OK	01:20	

Data Download Date/Time: 08/17/2018 11:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 07/21/2018 02:54:50

Diagnostics Check	OK	02:48
Air Blank	0.000	02:49
Control Test	0.079	02:49
Air Blank	0.000	02:49
Subject Sample # 1	0.174	02:50
Breath Volume	1.855	02:50
Air Blank	0.000	02:50
Air Blank	0.000	02:52
Subject Sample # 2	0.171	02:53
Breath Volume	1.625	02:53
Air Blank	0.000	02:53
Control Test	0.080	02:54
Air Blank	0.000	02:54
Diagnostics Check	OK	02:54

Data Download Date/Time: 08/17/2018 11:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: QUINN

DATE/TIME OF TEST: 07/21/2018 23:43:11

Diagnostics Check	OK	23:36
Air Blank	0.000	23:37
Control Test	0.080	23:37
Air Blank	0.000	23:38
Subject Sample # 1	0.060	23:38
Breath Volume	2.773	23:38
Air Blank	0.000	23:39
Air Blank	0.000	23:41
Subject Sample # 2	0.059	23:41
Breath Volume	3.089	23:41
Air Blank	0.000	23:42
Control Test	0.081	23:42
Air Blank	0.000	23:43
Diagnostics Check	OK	23:43

Data Download Date/Time: 08/17/2018 11:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: SKIPPER

DATE/TIME OF TEST: 07/22/2018 02:27:40

Diagnostics Check	OK	02:21
Air Blank	0.000	02:22
Control Test	0.080	02:22
Air Blank	0.000	02:22
Subject Sample # 1	0.182	02:23
Breath Volume	2.578	02:23
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample # 2	0.170	02:26
Breath Volume	2.109	02:26
Air Blank	0.000	02:26
Control Test	0.080	02:27
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

Data Download Date/Time: 08/17/2018 11:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HOOVER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 07/22/2018 03:10:58

Diagnostics Check	OK	03:05
Air Blank	0.000	03:05
Control Test	0.080	03:05
Air Blank	0.000	03:06
Subject Sample # 1	0.122	03:06
Breath Volume	2.457	03:06
Air Blank	0.000	03:07
Air Blank	0.000	03:09
Subject Sample # 2	0.124	03:09
Breath Volume	2.242	03:09
Air Blank	0.000	03:10
Control Test	0.080	03:10
Air Blank	0.000	03:10
Diagnostics Check	OK	03:10

Data Download Date/Time: 08/17/2018 11:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: TITH

DATE/TIME OF TEST: 07/22/2018 03:59:37

Diagnostics Check	OK	03:56	
Air Blank	0.000	03:57	
Control Test	0.080	03:57	
Air Blank	0.000	03:57	
Subject Sample	# 1	03:58	Subject Test Refused
Breath Volume	0.000	03:58	
Air Blank	0.000	03:58	
Control Test	0.080	03:59	
Air Blank	0.000	03:59	
Diagnostics Check	OK	03:59	

Data Download Date/Time: 08/17/2018 11:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AUGUSTINE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WILLIAMS,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 13:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:38

SUBJECT'S LAST NAME: SCHEIBELHUT

DATE/TIME OF TEST: 07/22/2018 18:05:29

Diagnostics Check	OK	17:59
Air Blank	0.000	17:59
Control Test	0.080	17:59
Air Blank	0.000	18:00
Subject Sample # 1	0.071	18:00
Breath Volume	1.355	18:00
Air Blank	0.000	18:01
Air Blank	0.000	18:03
Subject Sample # 2	0.072	18:04
Breath Volume	1.453	18:04
Air Blank	0.000	18:04
Control Test	0.080	18:04
Air Blank	0.000	18:05
Diagnostics Check	OK	18:05

Data Download Date/Time: 08/17/2018 11:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AUGUSTINE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:52

SUBJECT'S LAST NAME: LUKE

DATE/TIME OF TEST: 07/22/2018 20:19:31

Diagnostics Check	OK	20:13
Air Blank	0.000	20:13
Control Test	0.080	20:14
Air Blank	0.000	20:14
Subject Sample # 1	0.000	20:15
Breath Volume	1.316	20:15
Air Blank	0.000	20:15
Air Blank	0.000	20:17
Subject Sample # 2	0.000	20:18
Breath Volume	2.058	20:18
Air Blank	0.000	20:18
Control Test	0.080	20:18
Air Blank	0.000	20:19
Diagnostics Check	OK	20:19

Data Download Date/Time: 08/17/2018 11:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 07/23/2018 02:03:28

Diagnostics Check	OK	02:00	
Air Blank	0.000	02:00	
Control Test	0.080	02:01	
Air Blank	0.000	02:01	
Subject Sample	# 1	02:02	Subject Test Refused
Breath Volume	0.000	02:02	
Air Blank	0.000	02:02	
Control Test	0.081	02:02	
Air Blank	0.000	02:03	
Diagnostics Check	OK	02:03	

Data Download Date/Time: 08/17/2018 11:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: CASH

DATE/TIME OF TEST: 07/24/2018 00:07:12

Diagnostics Check	OK	00:01
Air Blank	0.000	00:01
Control Test	0.080	00:02
Air Blank	0.000	00:02
Subject Sample # 1	0.000	00:03
Breath Volume	1.398	00:03
Air Blank	0.000	00:03
Air Blank	0.000	00:05
Subject Sample # 2	0.000	00:05
Breath Volume	1.800	00:05
Air Blank	0.000	00:06
Control Test	0.080	00:06
Air Blank	0.000	00:07
Diagnostics Check	OK	00:07

Data Download Date/Time: 08/17/2018 11:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 07/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: BETO

DATE/TIME OF TEST: 07/24/2018 02:17:15

Diagnostics Check	OK	02:05	
Air Blank	0.000	02:06	
Control Test	0.080	02:06	
Air Blank	0.000	02:07	
Subject Sample	# 1 0.122	02:10	Volume Not Met
Breath Volume	0.226	02:10	
Air Blank	0.000	02:10	
Air Blank	0.000	02:12	
Subject Sample	# 2 0.095	02:15	Volume Not Met
Breath Volume	0.300	02:15	
Air Blank	0.000	02:16	
Control Test	0.079	02:16	
Air Blank	0.000	02:17	
Diagnostics Check	OK	02:17	

Data Download Date/Time: 08/17/2018 11:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KASSIS,F
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: HERDON

DATE/TIME OF TEST: 07/25/2018 02:47:55

Diagnostics Check	OK	02:41
Air Blank	0.000	02:42
Control Test	0.080	02:42
Air Blank	0.000	02:43
Subject Sample # 1	0.108	02:43
Breath Volume	1.234	02:43
Air Blank	0.000	02:44
Air Blank	0.000	02:45
Subject Sample # 2	0.109	02:46
Breath Volume	1.605	02:46
Air Blank	0.000	02:47
Control Test	0.081	02:47
Air Blank	0.000	02:47
Diagnostics Check	OK	02:47

Data Download Date/Time: 08/17/2018 11:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 07/26/2018 03:48:08

Diagnostics Check	OK	03:45	
Air Blank	0.000	03:45	
Control Test	0.080	03:46	
Air Blank	0.000	03:46	
Subject Sample	# 1	03:46	Subject Test Refused
Breath Volume	0.000	03:46	
Air Blank	0.000	03:47	
Control Test	0.080	03:47	
Air Blank	0.000	03:48	
Diagnostics Check	OK	03:48	

Data Download Date/Time: 08/17/2018 11:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DIAZ,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:14

SUBJECT'S LAST NAME: GARNSEY

DATE/TIME OF TEST: 07/26/2018 05:40:24

Diagnostics Check	OK	05:34
Air Blank	0.000	05:34
Control Test	0.079	05:34
Air Blank	0.000	05:35
Subject Sample # 1	0.087	05:35
Breath Volume	2.105	05:35
Air Blank	0.000	05:36
Air Blank	0.000	05:38
Subject Sample # 2	0.084	05:38
Breath Volume	1.429	05:38
Air Blank	0.000	05:39
Control Test	0.080	05:39
Air Blank	0.000	05:40
Diagnostics Check	OK	05:40

Data Download Date/Time: 08/17/2018 11:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HAWKINS,A
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 07/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: ROLLET

DATE/TIME OF TEST: 07/27/2018 02:19:47

Diagnostics Check	OK	02:13
Air Blank	0.000	02:14
Control Test	0.080	02:14
Air Blank	0.000	02:15
Subject Sample # 1	0.000	02:15
Breath Volume	2.253	02:15
Air Blank	0.000	02:16
Air Blank	0.000	02:18
Subject Sample # 2	0.000	02:18
Breath Volume	2.410	02:18
Air Blank	0.000	02:18
Control Test	0.080	02:19
Air Blank	0.000	02:19
Diagnostics Check	OK	02:19

Data Download Date/Time: 08/17/2018 11:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY, SHANON, E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BROWN, T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: BROCK

DATE/TIME OF TEST: 07/28/2018 23:46:31

Diagnostics Check	OK	23:37
Air Blank	0.000	23:38
Control Test	0.079	23:38
Air Blank	0.000	23:38
Subject Sample # 1	0.173	23:40
Breath Volume	1.144	23:40
Air Blank	0.000	23:41
Air Blank	0.000	23:43
Subject Sample # 2	0.168	23:44
Breath Volume	1.402	23:44
Air Blank	0.000	23:45
Control Test	0.079	23:45
Air Blank	0.000	23:46
Diagnostics Check	OK	23:46

Data Download Date/Time: 08/17/2018 11:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY, SHANON, E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SPLATT, R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: CARIAS TORRES

DATE/TIME OF TEST: 07/29/2018 05:27:53

Diagnostics Check	OK	05:16	
Air Blank	0.000	05:16	
Control Test	0.080	05:16	
Air Blank	0.000	05:17	
Subject Sample	# 1	05:20	No Sample Provided
Breath Volume	0.000	05:20	
Air Blank	0.000	05:21	
Air Blank	0.000	05:22	
Subject Sample	# 2	0.243	05:23
Breath Volume	1.230	05:23	
Air Blank	0.000	05:24	
Air Blank	0.000	05:25	
Subject Sample	# 3	0.232	05:26
Breath Volume	1.179	05:26	
Air Blank	0.000	05:26	
Control Test	0.079	05:27	
Air Blank	0.000	05:27	
Diagnostics Check	OK	05:27	

Data Download Date/Time: 08/17/2018 11:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001109

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,WILLIAM,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GREEN,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: VENEGAS

DATE/TIME OF TEST: 07/07/2018 02:59:02

Diagnostics Check	OK	02:49	
Air Blank	0.000	02:49	
Control Test	0.079	02:49	
Air Blank	0.000	02:50	
Subject Sample	# 1	02:53	No Sample Provided
Breath Volume	0.000	02:53	
Air Blank	0.000	02:54	
Air Blank	0.000	02:56	
Subject Sample	# 2	02:57	Subject Test Refused
Breath Volume	0.000	02:57	
Air Blank	0.000	02:58	
Control Test	0.080	02:58	
Air Blank	0.000	02:58	
Diagnostics Check	OK	02:59	

Data Download Date/Time: 07/23/2018 16:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001109

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,WILLIAM,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GREEN,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: DOUGLAS

DATE/TIME OF TEST: 07/07/2018 23:39:51

Diagnostics Check	OK	23:33
Air Blank	0.000	23:33
Control Test	0.079	23:34
Air Blank	0.000	23:34
Subject Sample # 1	0.082	23:35
Breath Volume	4.242	23:35
Air Blank	0.000	23:36
Air Blank	0.000	23:37
Subject Sample # 2	0.076	23:38
Breath Volume	3.445	23:38
Air Blank	0.000	23:38
Control Test	0.078	23:39
Air Blank	0.000	23:39
Diagnostics Check	OK	23:39

Data Download Date/Time: 07/23/2018 16:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001109

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARTNER,ARTHUR,F
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GARTNER,ARTHUR
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 20:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: HEMMINGWAY

DATE/TIME OF TEST: 07/19/2018 22:02:56

Diagnostics Check	OK	21:55
Air Blank	0.000	21:56
Control Test	0.077	21:56
Air Blank	0.000	21:57
Subject Sample # 1	0.132	21:58
Breath Volume	1.601	21:58
Air Blank	0.000	21:58
Air Blank	0.000	22:00
Subject Sample # 2	0.136	22:01
Breath Volume	1.468	22:01
Air Blank	0.000	22:02
Control Test	0.077	22:02
Air Blank	0.000	22:02
Diagnostics Check	OK	22:02

Data Download Date/Time: 07/23/2018 16:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001109

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARTNER,ARTHUR,F
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GARTNER,ARTHUR
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: CUBBISON

DATE/TIME OF TEST: 07/21/2018 02:47:49

Diagnostics Check	OK	02:41	
Air Blank	0.000	02:41	
Control Test	0.077	02:41	
Air Blank	0.000	02:42	
Subject Sample	# 1 0.000	02:43	
Breath Volume	2.546	02:43	
Air Blank	0.000	02:43	
Air Blank	0.000	02:45	
Subject Sample	# 2 0.000	02:46	
Breath Volume	2.906	02:46	
Air Blank	0.000	02:46	
Control Test	0.079	02:47	
Air Blank		02:47	RFI Detect
Air Blank	0.000	02:47	

Data Download Date/Time: 07/23/2018 16:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAILEY,RAYMOND,F
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BAILEY,RAYMOND
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 19:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:22

SUBJECT'S LAST NAME: HANEY

DATE/TIME OF TEST: 07/03/2018 20:48:48

Diagnostics Check	OK	20:42
Air Blank	0.000	20:42
Control Test	0.080	20:43
Air Blank	0.000	20:43
Subject Sample # 1	0.181	20:44
Breath Volume	2.832	20:44
Air Blank	0.000	20:44
Air Blank	0.000	20:46
Subject Sample # 2	0.177	20:47
Breath Volume	2.488	20:47
Air Blank	0.000	20:47
Control Test	0.080	20:48
Air Blank	0.000	20:48
Diagnostics Check	OK	20:48

Data Download Date/Time: 07/06/2018 19:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCNABB,KENNETH,E
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MCNABB,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: PRATT

DATE/TIME OF TEST: 07/05/2018 15:59:16

Diagnostics Check	OK	15:56	
Air Blank	0.000	15:56	
Control Test	0.080	15:57	
Air Blank	0.000	15:57	
Subject Sample	# 1	15:57	Subject Test Refused
Breath Volume	0.000	15:57	
Air Blank	0.000	15:58	
Control Test	0.081	15:58	
Air Blank	0.000	15:59	
Diagnostics Check	OK	15:59	

Data Download Date/Time: 07/06/2018 19:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAILEY,RAYMOND,F
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BAILEY,RAYMOND
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 18:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:16

SUBJECT'S LAST NAME: WILLIARD

DATE/TIME OF TEST: 07/07/2018 19:43:14

Diagnostics Check	OK	19:36
Air Blank	0.000	19:37
Control Test	0.081	19:37
Air Blank	0.000	19:38
Subject Sample # 1	0.153	19:38
Breath Volume	1.328	19:38
Air Blank	0.000	19:39
Air Blank	0.000	19:41
Subject Sample # 2	0.148	19:41
Breath Volume	1.300	19:41
Air Blank	0.000	19:42
Control Test	0.080	19:42
Air Blank	0.000	19:43
Diagnostics Check	OK	19:43

Data Download Date/Time: 08/03/2018 12:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWMAN,JOANNA,N
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BOWMAN,JOANNA
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 07/14/2018 15:01:45

Diagnostics Check	OK	14:57	
Air Blank	0.000	14:57	
Control Test	0.081	14:58	
Air Blank	0.000	14:58	
Subject Sample	# 1	15:01	Subject Test Refused
Breath Volume	0.000	15:01	
Air Blank		15:01	RFI Detect
Air Blank		15:01	RFI Detect

Data Download Date/Time: 08/03/2018 12:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GENTRY,JONATHAN,P
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: GENTRY,JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 07/22/2018 00:17:05

Diagnostics Check	OK	00:15	
Air Blank	0.000	00:15	
Control Test	0.081	00:16	
Air Blank		00:16	RFI Detect
Air Blank	0.000	00:17	

Data Download Date/Time: 08/03/2018 12:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GENTRY,JONATHAN,P
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: GENTRY,JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 07/22/2018 00:29:31

Diagnostics Check	OK	00:20	
Air Blank	0.000	00:21	
Control Test	0.082	00:21	
Air Blank	0.000	00:21	
Subject Sample	# 1	00:25	No Sample Provided
Breath Volume	0.000	00:25	
Air Blank	0.000	00:25	
Air Blank	0.000	00:27	
Subject Sample	# 2	00:28	Subject Test Refused
Breath Volume	0.000	00:28	
Air Blank	0.000	00:28	
Control Test	0.081	00:28	
Air Blank	0.000	00:29	
Diagnostics Check	OK	00:29	

Data Download Date/Time: 08/03/2018 12:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,RYAN,W
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BROWN,RYAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 15:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:55

SUBJECT'S LAST NAME: SISTRUNK

DATE/TIME OF TEST: 07/23/2018 16:46:21

Diagnostics Check	OK	16:40
Air Blank	0.000	16:41
Control Test	0.082	16:41
Air Blank	0.000	16:41
Subject Sample # 1	0.000	16:42
Breath Volume	3.406	16:42
Air Blank	0.000	16:42
Air Blank	0.000	16:44
Subject Sample # 2	0.000	16:45
Breath Volume	2.050	16:45
Air Blank	0.000	16:45
Control Test	0.081	16:45
Air Blank	0.000	16:46
Diagnostics Check	OK	16:46

Data Download Date/Time: 08/03/2018 12:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GENTRY,JONATHAN,P
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: GENTRY,JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: PENCE

DATE/TIME OF TEST: 07/25/2018 00:10:16

Diagnostics Check	OK	00:03
Air Blank	0.000	00:03
Control Test	0.082	00:03
Air Blank	0.000	00:04
Subject Sample # 1	0.157	00:05
Breath Volume	1.332	00:05
Air Blank	0.000	00:06
Air Blank	0.000	00:08
Subject Sample # 2	0.163	00:08
Breath Volume	1.164	00:08
Air Blank	0.000	00:09
Control Test	0.081	00:09
Air Blank	0.000	00:10
Diagnostics Check	OK	00:10

Data Download Date/Time: 08/03/2018 12:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001116

INSTRUMENT REGISTERED TO: EGLIN AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOSA,ROBERT,
BREATH TEST OPERATOR AGENCY: Eglin AFB PD
ARRESTING OFFICER: SOSA,ROBERT
ARRESTING OFFICER AGENCY: Eglin AFB PD
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: SNUFFY

DATE/TIME OF TEST: 07/07/2018 19:35:38

Diagnostics Check	OK	19:29
Air Blank	0.000	19:29
Control Test	0.079	19:29
Air Blank	0.000	19:30
Subject Sample # 1	0.000	19:30
Breath Volume	2.503	19:30
Air Blank	0.000	19:31
Air Blank	0.000	19:33
Subject Sample # 2	0.000	19:34
Breath Volume	2.128	19:34
Air Blank	0.000	19:34
Control Test	0.078	19:35
Air Blank	0.000	19:35
Diagnostics Check	OK	19:35

Data Download Date/Time: 08/10/2018 14:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001116

INSTRUMENT REGISTERED TO: EGLIN AFB

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: POWELL,RYAN,R

BREATH TEST OPERATOR AGENCY: Eglin AFB PD

ARRESTING OFFICER: DAVISON,CORNTNEY

ARRESTING OFFICER AGENCY: Eglin AFB PD

LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 05218080A

DRY GAS EXPIRY DATE: 05/05/2020

VIOLATION CODE: DUI

ARREST TIME: 00:55

GENDER: Female

OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: LYAKHOV

DATE/TIME OF TEST: 07/15/2018 01:36:57

Diagnostics Check	OK	01:28
Air Blank	0.000	01:29
Control Test	0.079	01:29
Air Blank	0.000	01:29
Subject Sample # 1	0.207	01:32
Breath Volume	1.453	01:32
Air Blank	0.000	01:32
Air Blank	0.000	01:34
Subject Sample # 2	0.195	01:35
Breath Volume	1.230	01:35
Air Blank	0.000	01:36
Control Test	0.078	01:36
Air Blank	0.000	01:36
Diagnostics Check	OK	01:36

Data Download Date/Time: 08/10/2018 14:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINTERS,LANE,A
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: REGAN,ERICK
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 07/01/2018 01:49:01

Diagnostics Check	OK	01:36
Air Blank	0.000	01:37
Control Test	0.082	01:37
Air Blank	0.000	01:37
Subject Sample # 1	0.178	01:39
Breath Volume	1.847	01:39
Air Blank	0.000	01:40
Air Blank	0.000	01:42
Subject Sample # 2	0.199	01:44
Breath Volume	1.632	01:44
Air Blank	0.000	01:45
Air Blank	0.000	01:46
Subject Sample # 3	0.195	01:47
Breath Volume	2.464	01:47
Air Blank	0.000	01:48
Control Test	0.080	01:48
Air Blank	0.000	01:48
Diagnostics Check	OK	01:49

Data Download Date/Time: 07/11/2018 14:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS, ADAM, M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: MORRIS, ADAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: KIRKWOOD

DATE/TIME OF TEST: 07/03/2018 00:46:31

Diagnostics Check	OK	00:40
Air Blank	0.000	00:40
Control Test	0.081	00:41
Air Blank	0.000	00:41
Subject Sample # 1	0.150	00:42
Breath Volume	2.718	00:42
Air Blank	0.000	00:42
Air Blank	0.000	00:44
Subject Sample # 2	0.151	00:45
Breath Volume	2.824	00:45
Air Blank	0.000	00:45
Control Test	0.082	00:45
Air Blank	0.000	00:46
Diagnostics Check	OK	00:46

Data Download Date/Time: 07/11/2018 14:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOTTINGHAM, CHRISTOPHER, R
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: NOTTINGHAM, CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 07:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:25

SUBJECT'S LAST NAME: IRIZARRY

DATE/TIME OF TEST: 07/03/2018 08:52:52

Diagnostics Check	OK	08:46
Air Blank	0.000	08:46
Control Test	0.082	08:47
Air Blank	0.000	08:47
Subject Sample # 1	0.000	08:48
Breath Volume	1.593	08:48
Air Blank	0.000	08:48
Air Blank	0.000	08:50
Subject Sample # 2	0.000	08:51
Breath Volume	1.835	08:51
Air Blank	0.000	08:51
Control Test	0.082	08:52
Air Blank	0.000	08:52
Diagnostics Check	OK	08:52

Data Download Date/Time: 07/11/2018 14:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRIOLAT,CASEY,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BRIOLAT,CASEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: GETZEN

DATE/TIME OF TEST: 07/06/2018 23:32:31

Diagnostics Check	OK	23:21	
Air Blank	0.000	23:21	
Control Test	0.081	23:21	
Air Blank	0.000	23:22	
Subject Sample	# 1	23:25	No Sample Provided
Breath Volume	0.000	23:25	
Air Blank	0.000	23:26	
Air Blank	0.000	23:28	
Subject Sample	# 2	23:31	No Sample Provided
Breath Volume	0.000	23:31	
Air Blank	0.000	23:31	
Control Test	0.080	23:31	
Air Blank	0.000	23:32	
Diagnostics Check	OK	23:32	

Data Download Date/Time: 07/11/2018 14:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BECKMAN,KEVIN,E
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: BECKMAN,KEVIN
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: GOVORUHK

DATE/TIME OF TEST: 07/08/2018 23:35:12

Diagnostics Check	OK	23:28
Air Blank	0.000	23:29
Control Test	0.080	23:29
Air Blank	0.000	23:29
Subject Sample # 1	0.153	23:30
Breath Volume	2.664	23:30
Air Blank	0.000	23:30
Air Blank	0.000	23:32
Subject Sample # 2	0.161	23:33
Breath Volume	3.625	23:33
Air Blank	0.000	23:34
Control Test	0.079	23:34
Air Blank	0.000	23:35
Diagnostics Check	OK	23:35

Data Download Date/Time: 07/11/2018 14:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOKENDORF,BRIAN,K
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DOKENDORF,BRIAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 13:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:30

SUBJECT'S LAST NAME: HENSEL

DATE/TIME OF TEST: 07/09/2018 14:55:57

Diagnostics Check	OK	14:50
Air Blank	0.000	14:50
Control Test	0.081	14:50
Air Blank	0.000	14:51
Subject Sample # 1	0.000	14:51
Breath Volume	2.257	14:51
Air Blank	0.000	14:52
Air Blank	0.000	14:54
Subject Sample # 2	0.000	14:54
Breath Volume	1.144	14:54
Air Blank	0.000	14:55
Control Test	0.081	14:55
Air Blank	0.000	14:55
Diagnostics Check	OK	14:55

Data Download Date/Time: 07/11/2018 14:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUCHANAN,SUMMER,N
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KELLY,WESLEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 09:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:15

SUBJECT'S LAST NAME: LOFTIS

DATE/TIME OF TEST: 07/11/2018 10:42:15

Diagnostics Check	OK	10:36
Air Blank	0.000	10:36
Control Test	0.081	10:37
Air Blank	0.000	10:37
Subject Sample # 1	0.000	10:37
Breath Volume	1.156	10:37
Air Blank	0.000	10:38
Air Blank	0.000	10:40
Subject Sample # 2	0.000	10:40
Breath Volume	2.449	10:40
Air Blank	0.000	10:41
Control Test	0.082	10:41
Air Blank	0.000	10:42
Diagnostics Check	OK	10:42

Data Download Date/Time: 07/11/2018 14:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARROW,M,A
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: LARKIN,J
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: ROBINETTE

DATE/TIME OF TEST: 07/11/2018 22:56:09

Diagnostics Check	OK	22:41	
Air Blank	0.000	22:42	
Control Test	0.082	22:42	
Air Blank	0.000	22:43	
Subject Sample	# 1 0.111	22:43	
Breath Volume	1.285	22:43	
Air Blank	0.000	22:44	
Air Blank	0.000	22:46	
Subject Sample	# 2 0.102	22:49	Volume Not Met
Breath Volume	0.753	22:49	
Air Blank	0.000	22:49	
Air Blank	0.000	22:51	
Subject Sample	# 3	22:54	Subject Test Refused
Breath Volume	0.000	22:54	
Air Blank	0.000	22:55	
Control Test	0.081	22:55	
Air Blank	0.000	22:56	
Diagnostics Check	OK	22:56	

Data Download Date/Time: 08/07/2018 17:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AIKEN,DURRIEL,R
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: AIKEN,DURRIEL
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 07/15/2018 02:16:20

Diagnostics Check	OK	02:09
Air Blank	0.000	02:09
Control Test	0.082	02:09
Air Blank	0.000	02:10
Subject Sample # 1	0.112	02:11
Breath Volume	1.664	02:11
Air Blank	0.000	02:12
Air Blank	0.000	02:14
Subject Sample # 2	0.119	02:14
Breath Volume	1.804	02:14
Air Blank	0.000	02:15
Control Test	0.082	02:15
Air Blank	0.000	02:16
Diagnostics Check	OK	02:16

Data Download Date/Time: 08/07/2018 17:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINTERS,LANE,A
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: DODD,L
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: KINSMAN JR

DATE/TIME OF TEST: 07/15/2018 03:45:02

Diagnostics Check	OK	03:38
Air Blank	0.000	03:38
Control Test	0.081	03:39
Air Blank	0.000	03:39
Subject Sample # 1	0.193	03:40
Breath Volume	1.988	03:40
Air Blank	0.000	03:40
Air Blank	0.000	03:42
Subject Sample # 2	0.193	03:43
Breath Volume	2.878	03:43
Air Blank	0.000	03:44
Control Test	0.082	03:44
Air Blank	0.000	03:44
Diagnostics Check	OK	03:45

Data Download Date/Time: 08/07/2018 17:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOTTINGHAM, CHRISTOPHER,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: NOTTINGHAM, CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 15:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: ADAMO

DATE/TIME OF TEST: 07/16/2018 17:23:36

Diagnostics Check	OK	17:17
Air Blank	0.000	17:18
Control Test	0.082	17:18
Air Blank	0.000	17:18
Subject Sample # 1	0.000	17:19
Breath Volume	3.101	17:19
Air Blank	0.000	17:19
Air Blank	0.000	17:21
Subject Sample # 2	0.000	17:22
Breath Volume	2.886	17:22
Air Blank	0.000	17:22
Control Test	0.082	17:23
Air Blank	0.000	17:23
Diagnostics Check	OK	17:23

Data Download Date/Time: 08/07/2018 17:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BECKMAN,KEVIN,E
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: IVEY,B
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: CHAINE

DATE/TIME OF TEST: 07/17/2018 01:09:47

Diagnostics Check	OK	01:00
Air Blank	0.000	01:01
Control Test	0.080	01:01
Air Blank	0.000	01:02
Subject Sample # 1	0.165	01:03
Breath Volume	1.308	01:03
Air Blank	0.000	01:04
Air Blank	0.000	01:06
Subject Sample # 2	0.171	01:08
Breath Volume	1.457	01:08
Air Blank	0.000	01:08
Control Test	0.080	01:09
Air Blank	0.000	01:09
Diagnostics Check	OK	01:09

Data Download Date/Time: 08/07/2018 17:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEASE, MATTHEW, W
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: JONES, RANDALL
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 12:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:18

SUBJECT'S LAST NAME: LAURO

DATE/TIME OF TEST: 07/20/2018 13:16:03

Diagnostics Check	OK	13:02	
Air Blank	0.000	13:02	
Control Test	0.081	13:03	
Air Blank	0.000	13:03	
Subject Sample	# 1 0.237	13:05	
Breath Volume	1.445	13:05	
Air Blank	0.000	13:06	
Air Blank	0.000	13:07	
Subject Sample	# 2 0.234	13:10	Volume Not Met
Breath Volume	0.000	13:10	
Air Blank	0.000	13:11	
Air Blank	0.000	13:13	
Subject Sample	# 3 0.232	13:14	
Breath Volume	1.652	13:14	
Air Blank	0.000	13:15	
Control Test	0.079	13:15	
Air Blank	0.000	13:15	
Diagnostics Check	OK	13:16	

Data Download Date/Time: 08/07/2018 17:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,BARRY,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: NIXON,BARRY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:23

SUBJECT'S LAST NAME: CHAMBLEY

DATE/TIME OF TEST: 07/20/2018 20:49:46

Diagnostics Check	OK	20:43
Air Blank	0.000	20:44
Control Test	0.081	20:44
Air Blank	0.000	20:45
Subject Sample # 1	0.000	20:45
Breath Volume	2.355	20:45
Air Blank	0.000	20:46
Air Blank	0.000	20:48
Subject Sample # 2	0.000	20:48
Breath Volume	2.402	20:48
Air Blank	0.000	20:48
Control Test	0.082	20:49
Air Blank	0.000	20:49
Diagnostics Check	OK	20:49

Data Download Date/Time: 08/07/2018 17:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,BARRY,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: WHITE,SEAN
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: OCONNELL

DATE/TIME OF TEST: 07/26/2018 00:01:55

Diagnostics Check	OK	23:56
Air Blank	0.000	23:56
Control Test	0.082	23:56
Air Blank	0.000	23:57
Subject Sample # 2	0.146	23:57
Breath Volume	1.718	23:57
Air Blank	0.000	23:58
Air Blank	0.000	00:00
Subject Sample # 1	0.153	00:00
Breath Volume	1.601	00:00
Air Blank	0.000	00:01
Control Test	0.082	00:01
Air Blank	0.000	00:01
Diagnostics Check	OK	00:01

Data Download Date/Time: 08/07/2018 17:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOTTINGHAM, CHRISTOPHER, R
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: NOTTINGHAM, CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 11:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: GAUS

DATE/TIME OF TEST: 07/26/2018 13:28:11

Diagnostics Check	OK	13:21
Air Blank	0.000	13:22
Control Test	0.081	13:22
Air Blank	0.000	13:23
Subject Sample # 1	0.000	13:23
Breath Volume	2.179	13:23
Air Blank	0.000	13:24
Air Blank	0.000	13:26
Subject Sample # 2	0.000	13:26
Breath Volume	1.636	13:26
Air Blank	0.000	13:27
Control Test	0.082	13:27
Air Blank	0.000	13:28
Diagnostics Check	OK	13:28

Data Download Date/Time: 08/07/2018 17:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS, ADAM, M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: MORRIS, ADAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: LEONHARDT

DATE/TIME OF TEST: 07/26/2018 23:59:17

Diagnostics Check	OK	23:52
Air Blank	0.000	23:53
Control Test	0.081	23:53
Air Blank	0.000	23:54
Subject Sample # 1	0.166	23:54
Breath Volume	1.328	23:54
Air Blank	0.000	23:55
Air Blank	0.000	23:57
Subject Sample # 2	0.163	23:57
Breath Volume	1.632	23:57
Air Blank	0.000	23:58
Control Test	0.082	23:58
Air Blank	0.000	23:59
Diagnostics Check	OK	23:59

Data Download Date/Time: 08/07/2018 17:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK, TYLER, J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: CORRALES, JOEL
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: BIDELSPACH

DATE/TIME OF TEST: 07/27/2018 23:43:35

Diagnostics Check	OK	23:36
Air Blank	0.000	23:37
Control Test	0.081	23:37
Air Blank	0.000	23:38
Subject Sample # 1	0.155	23:39
Breath Volume	1.136	23:39
Air Blank	0.000	23:39
Air Blank	0.000	23:41
Subject Sample # 2	0.158	23:42
Breath Volume	1.511	23:42
Air Blank	0.000	23:42
Control Test	0.079	23:42
Air Blank	0.000	23:43
Diagnostics Check	OK	23:43

Data Download Date/Time: 08/07/2018 17:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK, TYLER, J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: FRAGA, JACKSON
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: VAZQUEZ-CRUZ

DATE/TIME OF TEST: 07/28/2018 01:49:03

Diagnostics Check	OK	01:37	
Air Blank	0.000	01:37	
Control Test	0.082	01:37	
Air Blank	0.000	01:38	
Subject Sample	# 1 0.122	01:41	Volume Not Met
Breath Volume	0.785	01:41	
Air Blank	0.000	01:42	
Air Blank	0.000	01:43	
Subject Sample	# 2 0.129	01:44	
Breath Volume	2.195	01:44	
Air Blank	0.000	01:45	
Air Blank	0.000	01:46	
Subject Sample	# 3 0.129	01:47	
Breath Volume	1.871	01:47	
Air Blank	0.000	01:48	
Control Test	0.082	01:48	
Air Blank	0.000	01:48	
Diagnostics Check	OK	01:49	

Data Download Date/Time: 08/07/2018 17:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,RAEFORD,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GRIFFIN,RAEFORD
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: MILNE

DATE/TIME OF TEST: 07/29/2018 03:58:01

Diagnostics Check	OK	03:51
Air Blank	0.000	03:51
Control Test	0.081	03:52
Air Blank	0.000	03:52
Subject Sample # 1	0.027	03:53
Breath Volume	2.136	03:53
Air Blank	0.000	03:54
Air Blank	0.000	03:55
Subject Sample # 2	0.025	03:56
Breath Volume	1.519	03:56
Air Blank	0.000	03:57
Control Test	0.082	03:57
Air Blank	0.000	03:57
Diagnostics Check	OK	03:58

Data Download Date/Time: 08/07/2018 17:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLYTH,JACOB,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BLYTH,JACOB
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:21

SUBJECT'S LAST NAME: LINDSAY

DATE/TIME OF TEST: 07/29/2018 18:53:56

Diagnostics Check	OK	18:42	
Air Blank	0.000	18:42	
Control Test	0.081	18:42	
Air Blank	0.000	18:43	
Subject Sample	# 1 0.168	18:46	Volume Not Met
Breath Volume	0.757	18:46	
Air Blank	0.000	18:47	
Air Blank	0.000	18:48	
Subject Sample	# 2 0.153	18:49	
Breath Volume	1.984	18:49	
Air Blank	0.000	18:49	
Air Blank	0.000	18:51	
Subject Sample	# 3 0.165	18:52	
Breath Volume	2.414	18:52	
Air Blank	0.000	18:53	
Control Test	0.081	18:53	
Air Blank	0.000	18:53	
Diagnostics Check	OK	18:53	

Data Download Date/Time: 08/07/2018 17:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, WILLIAM, T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: BORNE

DATE/TIME OF TEST: 07/30/2018 01:23:54

Diagnostics Check	OK	01:12	
Air Blank	0.000	01:12	
Control Test	0.082	01:13	
Air Blank	0.000	01:13	
Subject Sample	# 1	01:16	No Sample Provided
Breath Volume	0.000	01:16	
Air Blank	0.000	01:17	
Air Blank	0.000	01:19	
Subject Sample	# 2	01:22	No Sample Provided
Breath Volume	0.000	01:22	
Air Blank	0.000	01:23	
Control Test	0.081	01:23	
Air Blank	0.000	01:23	
Diagnostics Check	OK	01:23	

Data Download Date/Time: 08/07/2018 17:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOSEPH,CAMERON,M
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: JOSEPH,CAMERON
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 437738
DRY GAS EXPIRY DATE: 06/08/2019
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: DREILING

DATE/TIME OF TEST: 07/01/2018 02:51:11

Diagnostics Check	OK	02:44
Air Blank	0.000	02:45
Control Test	0.081	02:45
Air Blank	0.000	02:46
Subject Sample # 1	0.000	02:46
Breath Volume	1.433	02:46
Air Blank	0.000	02:47
Air Blank	0.000	02:49
Subject Sample # 2	0.000	02:49
Breath Volume	1.148	02:49
Air Blank	0.000	02:50
Control Test	0.081	02:50
Air Blank	0.000	02:51
Diagnostics Check	OK	02:51

Data Download Date/Time: 07/06/2018 18:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001128

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRENCH, CALEB, L
BREATH TEST OPERATOR AGENCY: Putnam County SO
ARRESTING OFFICER: DAVIS, JARED
ARRESTING OFFICER AGENCY: Putnam County SO
LAST AGENCY INSPECTION: 07/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A3
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 08:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:10

SUBJECT'S LAST NAME: MALPHURS

DATE/TIME OF TEST: 07/03/2018 09:40:28

Diagnostics Check	OK	09:37	
Air Blank	0.000	09:38	
Control Test	0.078	09:38	
Air Blank	0.000	09:38	
Subject Sample	# 1	09:39	Subject Test Refused
Breath Volume	0.000	09:39	
Air Blank	0.000	09:39	
Control Test	0.078	09:39	
Air Blank	0.000	09:40	
Diagnostics Check	OK	09:40	

Data Download Date/Time: 08/02/2018 18:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STONE,JIMMIE,N
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: NEALIS,JACOB
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 11:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:15

SUBJECT'S LAST NAME: COOMBS

DATE/TIME OF TEST: 07/03/2018 12:46:20

Diagnostics Check	OK	12:39
Air Blank	0.000	12:40
Control Test	0.077	12:40
Air Blank	0.000	12:40
Subject Sample # 1	0.000	12:41
Breath Volume	1.769	12:41
Air Blank	0.000	12:41
Air Blank	0.000	12:43
Subject Sample # 2	0.000	12:44
Breath Volume	1.703	12:44
Air Blank	0.000	12:45
Control Test	0.077	12:45
Air Blank	0.000	12:46
Diagnostics Check	OK	12:46

Data Download Date/Time: 07/30/2018 12:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COKER,BRYAN,C
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: MYERS,MATTHEW
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 15:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:30

SUBJECT'S LAST NAME: CONWAY

DATE/TIME OF TEST: 07/04/2018 16:00:45

Diagnostics Check	OK	15:50	
Air Blank	0.000	15:51	
Control Test	0.076	15:51	
Air Blank	0.000	15:52	
Subject Sample	# 1 0.000	15:55	Volume Not Met
Breath Volume	0.488	15:55	
Air Blank	0.000	15:55	
Air Blank	0.000	15:57	
Subject Sample	# 2	15:59	Subject Test Refused
Breath Volume	0.000	15:59	
Air Blank	0.000	15:59	
Control Test	0.076	16:00	
Air Blank	0.000	16:00	
Diagnostics Check	OK	16:00	

Data Download Date/Time: 07/30/2018 12:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRYAN,STEVEN,A
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: PONDER,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 12:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: NEGRON

DATE/TIME OF TEST: 07/05/2018 01:55:21

Diagnostics Check	OK	01:48
Air Blank	0.000	01:49
Control Test	0.077	01:49
Air Blank	0.000	01:49
Subject Sample # 1	0.203	01:50
Breath Volume	2.484	01:50
Air Blank	0.000	01:51
Air Blank	0.000	01:53
Subject Sample # 2	0.187	01:53
Breath Volume	1.144	01:53
Air Blank	0.000	01:54
Control Test	0.077	01:54
Air Blank	0.000	01:55
Diagnostics Check	OK	01:55

Data Download Date/Time: 07/30/2018 12:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSARD,JOSHUA,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: WILLIAMS,NICK
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: PITTELKOW

DATE/TIME OF TEST: 07/17/2018 02:18:52

Diagnostics Check	OK	02:12
Air Blank	0.000	02:13
Control Test	0.076	02:13
Air Blank	0.000	02:14
Subject Sample # 1	0.000	02:14
Breath Volume	1.703	02:14
Air Blank	0.000	02:15
Air Blank	0.000	02:17
Subject Sample # 2	0.000	02:17
Breath Volume	1.203	02:17
Air Blank	0.000	02:17
Control Test	0.077	02:18
Air Blank	0.000	02:18
Diagnostics Check	OK	02:18

Data Download Date/Time: 07/30/2018 12:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSARD,JOSHUA,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: VANZO,MICHAEL
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: ORTIZ TORRES

DATE/TIME OF TEST: 07/21/2018 02:24:37

Diagnostics Check	OK	02:16
Air Blank	0.000	02:17
Control Test	0.076	02:17
Air Blank	0.000	02:18
Subject Sample # 1	0.153	02:18
Breath Volume	1.390	02:18
Air Blank	0.000	02:19
Air Blank	0.000	02:21
Subject Sample # 2	0.153	02:23
Breath Volume	1.828	02:23
Air Blank	0.000	02:23
Control Test	0.077	02:24
Air Blank	0.000	02:24
Diagnostics Check	OK	02:24

Data Download Date/Time: 07/30/2018 12:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRYAN,STEVEN,A
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: GODENZI,CAMERON
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: YELVINGTON

DATE/TIME OF TEST: 07/28/2018 02:41:52

Diagnostics Check	OK	02:34	
Air Blank	0.000	02:34	
Control Test	0.077	02:35	
Air Blank	0.000	02:35	
Subject Sample	# 1 0.242	02:37	
Breath Volume	1.441	02:37	
Air Blank	0.000	02:37	
Air Blank	0.000	02:39	
Subject Sample	# 2 0.239	02:40	
Breath Volume	1.593	02:40	
Air Blank	0.000	02:40	
Control Test	0.074	02:41	Control Outside Tolerance
Air Blank	0.000	02:41	

Data Download Date/Time: 07/30/2018 12:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRYAN,STEVEN,A
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: GODENZI,CAMERON
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: YELVINGTON

DATE/TIME OF TEST: 07/28/2018 02:54:48

Diagnostics Check	OK	02:46
Air Blank	0.000	02:47
Control Test	0.076	02:47
Air Blank	0.000	02:47
Subject Sample # 1	0.233	02:48
Breath Volume	1.203	02:48
Air Blank	0.000	02:49
Air Blank	0.000	02:51
Subject Sample # 2	0.229	02:53
Breath Volume	1.449	02:53
Air Blank	0.000	02:53
Control Test	0.076	02:54
Air Blank	0.000	02:54
Diagnostics Check	OK	02:54

Data Download Date/Time: 07/30/2018 12:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRYAN,STEVEN,A
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: JOHANSMEYER,JONATHAN
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: WODZ

DATE/TIME OF TEST: 07/28/2018 03:19:13

Diagnostics Check	OK	03:12
Air Blank	0.000	03:12
Control Test	0.076	03:12
Air Blank	0.000	03:13
Subject Sample # 1	0.091	03:14
Breath Volume	2.562	03:14
Air Blank	0.000	03:15
Air Blank	0.000	03:16
Subject Sample # 2	0.082	03:17
Breath Volume	2.085	03:17
Air Blank	0.000	03:18
Control Test	0.076	03:18
Air Blank	0.000	03:19
Diagnostics Check	OK	03:19

Data Download Date/Time: 07/30/2018 12:40