

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DIXON, WESLEY, B

BREATH TEST OPERATOR AGENCY: FHP Troop H

ARRESTING OFFICER: DIXON, WESLEY

ARRESTING OFFICER AGENCY: FHP Troop H

LAST AGENCY INSPECTION: 08/15/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 29017080A3

DRY GAS EXPIRY DATE: 12/05/2019

VIOLATION CODE: DUI

ARREST TIME: 01:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 09/03/2018 02:48:17

Diagnostics Check	OK	02:47	
Air Blank	0.000	02:47	
Control Test	0.000	02:48	RFI Detect
Air Blank		02:48	RFI Detect

Data Download Date/Time: 09/20/2018 10:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DIXON, WESLEY, B

BREATH TEST OPERATOR AGENCY: FHP Troop H

ARRESTING OFFICER: DIXON, WESLEY

ARRESTING OFFICER AGENCY: FHP Troop H

LAST AGENCY INSPECTION: 08/15/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 29017080A3

DRY GAS EXPIRY DATE: 12/05/2019

VIOLATION CODE: DUI

ARREST TIME: 01:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 09/03/2018 02:53:13

Diagnostics Check	OK	02:50	
Air Blank	0.000	02:51	
Control Test	0.080	02:51	
Air Blank	0.000	02:52	
Subject Sample	# 1	02:52	RFI Detect
Breath Volume	0.000	02:52	
Air Blank		02:52	RFI Detect
Air Blank		02:53	RFI Detect

Data Download Date/Time: 09/20/2018 10:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DIXON, WESLEY, B

BREATH TEST OPERATOR AGENCY: FHP Troop H

ARRESTING OFFICER: DIXON, WESLEY

ARRESTING OFFICER AGENCY: FHP Troop H

LAST AGENCY INSPECTION: 08/15/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 29017080A3

DRY GAS EXPIRY DATE: 12/05/2019

VIOLATION CODE: DUI

ARREST TIME: 01:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 09/03/2018 02:56:13

Diagnostics Check	OK	02:54	
Air Blank	0.000	02:55	
Control Test	0.080	02:55	RFI Detect
Air Blank	0.000	02:56	

Data Download Date/Time: 09/20/2018 10:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DIXON, WESLEY, B

BREATH TEST OPERATOR AGENCY: FHP Troop H

ARRESTING OFFICER: DIXON, WESLEY

ARRESTING OFFICER AGENCY: FHP Troop H

LAST AGENCY INSPECTION: 08/15/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 29017080A3

DRY GAS EXPIRY DATE: 12/05/2019

VIOLATION CODE: DUI

ARREST TIME: 01:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 09/03/2018 03:06:06

Diagnostics Check	OK	02:58
Air Blank	0.000	02:59
Control Test	0.080	02:59
Air Blank	0.000	03:00
Subject Sample # 1	0.121	03:01
Breath Volume	3.359	03:01
Air Blank	0.000	03:02
Air Blank	0.000	03:04
Subject Sample # 2	0.114	03:04
Breath Volume	1.746	03:04
Air Blank	0.000	03:05
Control Test	0.078	03:05
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

Data Download Date/Time: 09/20/2018 10:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTHWAY,GEORGIA,K
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: CROUCH,CLIFF
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 08/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A3
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:48

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 09/03/2018 07:17:44

Diagnostics Check	OK	07:09
Air Blank	0.000	07:10
Control Test	0.080	07:10
Air Blank	0.000	07:11
Subject Sample # 1	0.027	07:13
Breath Volume	1.207	07:13
Air Blank	0.000	07:13
Air Blank	0.000	07:15
Subject Sample # 2	0.028	07:16
Breath Volume	1.132	07:16
Air Blank	0.000	07:16
Control Test	0.079	07:17
Air Blank	0.000	07:17
Diagnostics Check	OK	07:17

Data Download Date/Time: 09/20/2018 10:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BULLOCK,JOSHUA,R

BREATH TEST OPERATOR AGENCY: Tallahassee PD

ARRESTING OFFICER: HARVEY,R

ARRESTING OFFICER AGENCY: Tallahassee PD

LAST AGENCY INSPECTION: 08/15/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 29017080A3

DRY GAS EXPIRY DATE: 12/05/2019

VIOLATION CODE: DUI

ARREST TIME: 04:10

GENDER: Male

OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 09/06/2018 05:24:00

Diagnostics Check	OK	05:20	
Air Blank	0.000	05:21	
Control Test	0.080	05:21	
Air Blank	0.000	05:22	
Subject Sample	# 1	05:22	Subject Test Refused
Breath Volume	0.000	05:22	
Air Blank	0.000	05:23	
Control Test	0.080	05:23	
Air Blank	0.000	05:23	
Diagnostics Check	OK	05:24	

Data Download Date/Time: 09/20/2018 10:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DIXON, WESLEY, B

BREATH TEST OPERATOR AGENCY: FHP Troop H

ARRESTING OFFICER: DIXON, WESLEY

ARRESTING OFFICER AGENCY: FHP Troop H

LAST AGENCY INSPECTION: 08/15/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 29017080A3

DRY GAS EXPIRY DATE: 12/05/2019

VIOLATION CODE: DUI

ARREST TIME: 20:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: STUDDARD

DATE/TIME OF TEST: 09/06/2018 22:04:25

Diagnostics Check	OK	21:58
Air Blank	0.000	21:58
Control Test	0.080	21:58
Air Blank	0.000	21:59
Subject Sample # 1	0.000	21:59
Breath Volume	4.261	21:59
Air Blank	0.000	22:00
Air Blank	0.000	22:02
Subject Sample # 2	0.000	22:03
Breath Volume	4.527	22:03
Air Blank	0.000	22:03
Control Test	0.080	22:03
Air Blank	0.000	22:04
Diagnostics Check	OK	22:04

Data Download Date/Time: 09/20/2018 10:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEELE,R,L
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: STEELE,R
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 08/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A3
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 09/10/2018 06:31:24

Diagnostics Check	OK	06:25	
Air Blank	0.000	06:25	
Control Test	0.080	06:26	
Air Blank	0.000	06:26	
Subject Sample	# 1 0.168	06:27	
Breath Volume	1.402	06:27	
Air Blank	0.000	06:27	
Air Blank	0.000	06:29	
Subject Sample	# 2 0.169	06:30	
Breath Volume	1.652	06:30	
Air Blank		06:30	RFI Detect
Air Blank	0.000	06:31	

Data Download Date/Time: 09/20/2018 10:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PILKANIS,MARY,C

BREATH TEST OPERATOR AGENCY: Leon County SO

ARRESTING OFFICER: PILKANIS,MARY

ARRESTING OFFICER AGENCY: Leon County SO

LAST AGENCY INSPECTION: 08/15/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 29017080A3

DRY GAS EXPIRY DATE: 12/05/2019

VIOLATION CODE: DUI

ARREST TIME: 23:04

GENDER: Female

OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 09/13/2018 00:44:23

Diagnostics Check	OK	00:41	
Air Blank	0.000	00:42	
Control Test	0.080	00:42	
Air Blank	0.000	00:42	
Subject Sample	# 1	00:43	Subject Test Refused
Breath Volume	0.000	00:43	
Air Blank	0.000	00:43	
Control Test	0.080	00:43	
Air Blank	0.000	00:44	
Diagnostics Check	OK	00:44	

Data Download Date/Time: 09/20/2018 10:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PILKANIS,MARY,C

BREATH TEST OPERATOR AGENCY: Leon County SO

ARRESTING OFFICER: FLOYD,JEVON

ARRESTING OFFICER AGENCY: Leon County SO

LAST AGENCY INSPECTION: 08/15/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 29017080A3

DRY GAS EXPIRY DATE: 12/05/2019

VIOLATION CODE: DUI

ARREST TIME: 23:12

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: PITTMAN

DATE/TIME OF TEST: 09/13/2018 00:49:12

Diagnostics Check	OK	00:46	
Air Blank	0.000	00:46	
Control Test	0.080	00:47	
Air Blank	0.000	00:47	
Subject Sample	# 1	00:47	Subject Test Refused
Breath Volume	0.000	00:47	
Air Blank	0.000	00:48	
Control Test	0.080	00:48	
Air Blank	0.000	00:49	
Diagnostics Check	OK	00:49	

Data Download Date/Time: 09/20/2018 10:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: THOMPSON,JEFFREY,A

BREATH TEST OPERATOR AGENCY: Tallahassee PD

ARRESTING OFFICER: THOMPSON,JEFFREY

ARRESTING OFFICER AGENCY: Tallahassee PD

LAST AGENCY INSPECTION: 09/20/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 29017080A3

DRY GAS EXPIRY DATE: 12/05/2019

VIOLATION CODE: DUI

ARREST TIME: 22:31

GENDER: Female

OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: GRACE

DATE/TIME OF TEST: 09/21/2018 23:40:27

Diagnostics Check	OK	23:34
Air Blank	0.000	23:34
Control Test	0.080	23:34
Air Blank	0.000	23:35
Subject Sample # 1	0.203	23:36
Breath Volume	1.660	23:36
Air Blank	0.000	23:36
Air Blank	0.000	23:38
Subject Sample # 2	0.201	23:38
Breath Volume	1.609	23:38
Air Blank	0.000	23:39
Control Test	0.080	23:39
Air Blank	0.000	23:40
Diagnostics Check	OK	23:40

Data Download Date/Time: 10/25/2018 13:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,ANDREW,L
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: WESTMAN,PATRICK
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 09/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A3
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: XXXXXXXXXX

DATE/TIME OF TEST: 09/23/2018 00:04:52

Diagnostics Check	OK	23:58
Air Blank	0.000	23:58
Control Test	0.080	23:59
Air Blank	0.000	23:59
Subject Sample # 1	0.055	00:00
Breath Volume	3.367	00:00
Air Blank	0.000	00:00
Air Blank	0.000	00:02
Subject Sample # 2	0.057	00:03
Breath Volume	3.187	00:03
Air Blank	0.000	00:03
Control Test	0.079	00:04
Air Blank	0.000	00:04
Diagnostics Check	OK	00:04

Data Download Date/Time: 10/25/2018 13:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LECKINGER, MICHAEL, E

BREATH TEST OPERATOR AGENCY: FHP Troop H

ARRESTING OFFICER: CLEMENT, JOSEPH

ARRESTING OFFICER AGENCY: FHP Troop H

LAST AGENCY INSPECTION: 09/20/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 29017080A3

DRY GAS EXPIRY DATE: 12/05/2019

VIOLATION CODE: DUI

ARREST TIME: 01:52

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 09/28/2018 03:47:14

Diagnostics Check	OK	03:40
Air Blank	0.000	03:41
Control Test	0.080	03:41
Air Blank	0.000	03:42
Subject Sample # 1	0.127	03:42
Breath Volume	1.820	03:42
Air Blank	0.000	03:43
Air Blank	0.000	03:45
Subject Sample # 2	0.135	03:45
Breath Volume	3.058	03:45
Air Blank	0.000	03:46
Control Test	0.079	03:46
Air Blank	0.000	03:47
Diagnostics Check	OK	03:47

Data Download Date/Time: 10/25/2018 13:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLETCHER,ERIC,D
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: FLETCHER,ERIC
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 09/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A3
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: HATCHER

DATE/TIME OF TEST: 09/29/2018 23:30:22

Diagnostics Check	OK	23:23
Air Blank	0.000	23:24
Control Test	0.079	23:24
Air Blank	0.000	23:25
Subject Sample # 1	0.249	23:26
Breath Volume	3.882	23:26
Air Blank	0.000	23:26
Air Blank	0.000	23:28
Subject Sample # 2	0.259	23:28
Breath Volume	2.019	23:28
Air Blank	0.000	23:29
Control Test	0.077	23:29
Air Blank	0.000	23:30
Diagnostics Check	OK	23:30

Data Download Date/Time: 10/25/2018 13:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLER,DAVID,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: KELLER,DAVID
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 08/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: NANIC

DATE/TIME OF TEST: 09/01/2018 01:24:20

Diagnostics Check	OK	01:18
Air Blank	0.000	01:18
Control Test	0.079	01:18
Air Blank	0.000	01:19
Subject Sample # 1	0.124	01:19
Breath Volume	1.996	01:19
Air Blank	0.000	01:20
Air Blank	0.000	01:22
Subject Sample # 2	0.116	01:22
Breath Volume	1.269	01:22
Air Blank	0.000	01:23
Control Test	0.079	01:23
Air Blank	0.000	01:24
Diagnostics Check	OK	01:24

Data Download Date/Time: 09/20/2018 10:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GONZALEZ,AUSTIN,L

BREATH TEST OPERATOR AGENCY: Tallahassee PD

ARRESTING OFFICER: GONZALEZ,AUSTIN

ARRESTING OFFICER AGENCY: Tallahassee PD

LAST AGENCY INSPECTION: 08/09/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 06717080A6

DRY GAS EXPIRY DATE: 05/05/2019

VIOLATION CODE: DUI

ARREST TIME: 22:26

GENDER: Female

OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 09/15/2018 23:50:49

Diagnostics Check	OK	23:39	
Air Blank	0.000	23:39	
Control Test	0.079	23:40	
Air Blank	0.000	23:40	
Subject Sample	# 1 0.118	23:43	Volume Not Met
Breath Volume	0.082	23:43	
Air Blank	0.000	23:44	
Air Blank	0.000	23:46	
Subject Sample	# 2 0.125	23:49	Volume Not Met
Breath Volume	0.062	23:49	
Air Blank	0.000	23:49	
Control Test	0.078	23:50	
Air Blank	0.000	23:50	
Diagnostics Check	OK	23:50	

Data Download Date/Time: 09/20/2018 10:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARWOOD,LEE,M
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: HUNTER,ROBERT
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 09/20/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: GANTT

DATE/TIME OF TEST: 09/30/2018 04:39:29

Diagnostics Check	OK	04:26	
Air Blank	0.000	04:27	
Control Test	0.079	04:27	
Air Blank	0.000	04:27	
Subject Sample	# 1 0.159	04:31	Volume Not Met
Breath Volume	0.863	04:31	
Air Blank	0.000	04:31	
Air Blank	0.000	04:33	
Subject Sample	# 2 0.169	04:34	
Breath Volume	1.234	04:34	
Air Blank	0.000	04:35	
Air Blank	0.000	04:36	
Subject Sample	# 3 0.169	04:38	
Breath Volume	1.585	04:38	
Air Blank	0.000	04:38	
Control Test	0.078	04:38	
Air Blank	0.000	04:39	
Diagnostics Check	OK	04:39	

Data Download Date/Time: 10/25/2018 13:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORBIN,CLEVE,B
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: HARRISON,BOBBY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: OSVALDO

DATE/TIME OF TEST: 09/04/2018 00:38:39

Diagnostics Check	OK	00:32
Air Blank	0.000	00:32
Control Test	0.081	00:32
Air Blank	0.000	00:33
Subject Sample # 1	0.163	00:34
Breath Volume	5.390	00:34
Air Blank	0.000	00:34
Air Blank	0.000	00:36
Subject Sample # 2	0.149	00:37
Breath Volume	2.476	00:37
Air Blank	0.000	00:37
Control Test	0.080	00:38
Air Blank	0.000	00:38
Diagnostics Check	OK	00:38

Data Download Date/Time: 09/28/2018 10:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLAUSNER,ALEXANDRA,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: CLAUSNER,ALEXANDRA
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 16:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: MCGREGOR

DATE/TIME OF TEST: 09/07/2018 17:42:06

Diagnostics Check	OK	17:36
Air Blank	0.000	17:36
Control Test	0.079	17:36
Air Blank	0.000	17:37
Subject Sample # 1	0.000	17:37
Breath Volume	3.039	17:37
Air Blank	0.000	17:38
Air Blank	0.000	17:40
Subject Sample # 2	0.000	17:40
Breath Volume	3.160	17:40
Air Blank	0.000	17:41
Control Test	0.080	17:41
Air Blank	0.000	17:42
Diagnostics Check	OK	17:42

Data Download Date/Time: 09/28/2018 10:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLAUSNER,ALEXANDRA,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: CLAUSNER,ALEXANDRA
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 07:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:33

SUBJECT'S LAST NAME: SOTO-PANTOJA

DATE/TIME OF TEST: 09/09/2018 08:59:57

Diagnostics Check	OK	08:53
Air Blank	0.000	08:54
Control Test	0.080	08:54
Air Blank	0.000	08:55
Subject Sample # 1	0.123	08:55
Breath Volume	2.382	08:55
Air Blank	0.000	08:56
Air Blank	0.000	08:57
Subject Sample # 2	0.133	08:58
Breath Volume	3.113	08:58
Air Blank	0.000	08:59
Control Test	0.079	08:59
Air Blank	0.000	08:59
Diagnostics Check	OK	08:59

Data Download Date/Time: 09/28/2018 10:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000742

INSTRUMENT REGISTERED TO: VALPARAISO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIDSON, JEREMIAH, D
BREATH TEST OPERATOR AGENCY: Valparaiso PD
ARRESTING OFFICER: SOWELL, JONATHAN
ARRESTING OFFICER AGENCY: Eglin AFB PD
LAST AGENCY INSPECTION: 08/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: BARTO

DATE/TIME OF TEST: 09/16/2018 01:46:55

Diagnostics Check	OK	01:44	
Air Blank	0.000	01:44	
Control Test	0.080	01:44	
Air Blank	0.000	01:45	
Subject Sample	# 1	01:45	Subject Test Refused
Breath Volume	0.000	01:45	
Air Blank	0.000	01:46	
Control Test	0.080	01:46	
Air Blank	0.000	01:46	
Diagnostics Check	OK	01:46	

Data Download Date/Time: 09/29/2018 23:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000742

INSTRUMENT REGISTERED TO: VALPARAISO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIDSON, JEREMIAH, D
BREATH TEST OPERATOR AGENCY: Valparaiso PD
ARRESTING OFFICER: SOWELL, JONATHAN
ARRESTING OFFICER AGENCY: Eglin AFB PD
LAST AGENCY INSPECTION: 08/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: BARTO

DATE/TIME OF TEST: 09/16/2018 01:52:46

Diagnostics Check	OK	01:49	
Air Blank	0.000	01:50	
Control Test	0.080	01:50	
Air Blank	0.000	01:51	
Subject Sample	# 1	01:51	Subject Test Refused
Breath Volume	0.000	01:51	
Air Blank	0.000	01:51	
Control Test	0.080	01:52	
Air Blank	0.000	01:52	
Diagnostics Check	OK	01:52	

Data Download Date/Time: 09/29/2018 23:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000742

INSTRUMENT REGISTERED TO: VALPARAISO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIDSON, JEREMIAH, D
BREATH TEST OPERATOR AGENCY: Valparaiso PD
ARRESTING OFFICER: DE MARIA, DAVID
ARRESTING OFFICER AGENCY: Eglin AFB PD
LAST AGENCY INSPECTION: 08/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: GUERRERO

DATE/TIME OF TEST: 09/16/2018 02:04:06

Diagnostics Check	OK	01:56
Air Blank	0.000	01:56
Control Test	0.081	01:56
Air Blank	0.000	01:57
Subject Sample # 1	0.142	01:59
Breath Volume	1.597	01:59
Air Blank	0.000	01:59
Air Blank	0.000	02:01
Subject Sample # 2	0.154	02:02
Breath Volume	1.640	02:02
Air Blank	0.000	02:03
Control Test	0.078	02:03
Air Blank	0.000	02:04
Diagnostics Check	OK	02:04

Data Download Date/Time: 09/29/2018 23:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROOP,CHAD,E
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: ROOP,CHAD
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 08/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:02

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 09/01/2018 05:30:24

Diagnostics Check	OK	05:22
Air Blank	0.000	05:22
Control Test	0.078	05:23
Air Blank	0.000	05:23
Subject Sample # 1	0.181	05:26
Breath Volume	1.687	05:26
Air Blank	0.000	05:26
Air Blank	0.000	05:28
Subject Sample # 2	0.182	05:28
Breath Volume	1.312	05:28
Air Blank	0.000	05:29
Control Test	0.078	05:29
Air Blank	0.000	05:30
Diagnostics Check	OK	05:30

Data Download Date/Time: 09/26/2018 07:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAFKJEN,THOMAS,D
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: GAFKJEN,THOMAS
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 08/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 06:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: TURNAGE

DATE/TIME OF TEST: 09/01/2018 06:21:15

Diagnostics Check	OK	06:14
Air Blank	0.000	06:14
Control Test	0.078	06:14
Air Blank	0.000	06:15
Subject Sample # 1	0.183	06:16
Breath Volume	1.894	06:16
Air Blank	0.000	06:17
Air Blank	0.000	06:18
Subject Sample # 2	0.199	06:19
Breath Volume	1.417	06:19
Air Blank	0.000	06:20
Control Test	0.076	06:20
Air Blank	0.000	06:21
Diagnostics Check	OK	06:21

Data Download Date/Time: 09/26/2018 07:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENLEY, WILLIAM, C
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: HENLEY, WILLIAM
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 08/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 09/02/2018 02:41:51

Diagnostics Check	OK	02:35
Air Blank	0.000	02:35
Control Test	0.078	02:36
Air Blank	0.000	02:36
Subject Sample # 1	0.110	02:37
Breath Volume	1.906	02:37
Air Blank	0.000	02:37
Air Blank	0.000	02:39
Subject Sample # 2	0.118	02:40
Breath Volume	3.316	02:40
Air Blank	0.000	02:40
Control Test	0.079	02:41
Air Blank	0.000	02:41
Diagnostics Check	OK	02:41

Data Download Date/Time: 09/26/2018 07:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARACLE,MATTHEW,D
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: MARACLE,MATTHEW
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 08/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 06:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:30

SUBJECT'S LAST NAME: ALHAZZAA

DATE/TIME OF TEST: 09/02/2018 07:03:19

Diagnostics Check	OK	06:50	
Air Blank	0.000	06:51	
Control Test	0.078	06:51	
Air Blank	0.000	06:52	
Subject Sample	# 1 0.151	06:55	Volume Not Met
Breath Volume	0.441	06:55	
Air Blank	0.000	06:55	
Air Blank	0.000	06:57	
Subject Sample	# 2 0.145	06:58	
Breath Volume	1.968	06:58	
Air Blank	0.000	06:58	
Air Blank	0.000	07:00	
Subject Sample	# 3 0.144	07:01	
Breath Volume	1.984	07:01	
Air Blank	0.000	07:02	
Control Test	0.076	07:02	
Air Blank	0.000	07:03	
Diagnostics Check	OK	07:03	

Data Download Date/Time: 09/26/2018 07:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DURHAM,JAMES,S
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: DURHAM,JAMES
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: TUCKER

DATE/TIME OF TEST: 09/03/2018 17:58:05

Diagnostics Check	OK	17:52
Air Blank	0.000	17:52
Control Test	0.078	17:52
Air Blank	0.000	17:53
Subject Sample # 1	0.000	17:53
Breath Volume	2.460	17:53
Air Blank	0.000	17:54
Air Blank	0.000	17:56
Subject Sample # 2	0.000	17:56
Breath Volume	2.777	17:56
Air Blank	0.000	17:57
Control Test	0.079	17:57
Air Blank	0.000	17:58
Diagnostics Check	OK	17:58

Data Download Date/Time: 09/26/2018 07:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELLEY,SAMUEL,L
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: SHELLEY,SAMUEL
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 08/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: PELEZO

DATE/TIME OF TEST: 09/05/2018 23:12:43

Diagnostics Check	OK	23:06
Air Blank	0.000	23:06
Control Test	0.079	23:06
Air Blank	0.000	23:07
Subject Sample # 1	0.216	23:08
Breath Volume	2.261	23:08
Air Blank	0.000	23:08
Air Blank	0.000	23:10
Subject Sample # 2	0.211	23:11
Breath Volume	2.398	23:11
Air Blank	0.000	23:11
Control Test	0.078	23:12
Air Blank	0.000	23:12
Diagnostics Check	OK	23:12

Data Download Date/Time: 09/26/2018 07:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMPTON,L,I
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: HAMPTON,L
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 08/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: MONISH

DATE/TIME OF TEST: 09/06/2018 23:22:51

Diagnostics Check	OK	23:14
Air Blank	0.000	23:14
Control Test	0.078	23:14
Air Blank	0.000	23:15
Subject Sample # 1	0.045	23:17
Breath Volume	1.363	23:17
Air Blank	0.000	23:18
Air Blank	0.000	23:20
Subject Sample # 2	0.039	23:21
Breath Volume	1.230	23:21
Air Blank	0.000	23:21
Control Test	0.076	23:22
Air Blank	0.000	23:22
Diagnostics Check	OK	23:22

Data Download Date/Time: 09/26/2018 07:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEITE,STEVEN,G
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: LEITE,STEVEN
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 08/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: CROWSON

DATE/TIME OF TEST: 09/07/2018 00:25:47

Diagnostics Check	OK	00:19
Air Blank	0.000	00:19
Control Test	0.077	00:20
Air Blank	0.000	00:20
Subject Sample # 1	0.174	00:21
Breath Volume	1.375	00:21
Air Blank	0.000	00:21
Air Blank	0.000	00:23
Subject Sample # 2	0.166	00:24
Breath Volume	1.410	00:24
Air Blank	0.000	00:24
Control Test	0.078	00:25
Air Blank	0.000	00:25
Diagnostics Check	OK	00:25

Data Download Date/Time: 09/26/2018 07:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENLEY, WILLIAM, C
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: HENLEY, WILLIAM
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 08/21/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:43

SUBJECT'S LAST NAME: DAILEY

DATE/TIME OF TEST: 09/07/2018 23:11:18

Diagnostics Check	OK	23:04
Air Blank	0.000	23:04
Control Test	0.078	23:04
Air Blank	0.000	23:05
Subject Sample # 1	0.084	23:06
Breath Volume	2.070	23:06
Air Blank	0.000	23:07
Air Blank	0.000	23:09
Subject Sample # 2	0.078	23:09
Breath Volume	1.914	23:09
Air Blank	0.000	23:10
Control Test	0.077	23:10
Air Blank	0.000	23:11
Diagnostics Check	OK	23:11

Data Download Date/Time: 09/26/2018 07:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BYRD,BRANDON,C
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: WILLIAMS,BRANDON
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 08/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: CHATTIN

DATE/TIME OF TEST: 09/08/2018 20:02:10

Diagnostics Check	OK	19:55
Air Blank	0.000	19:55
Control Test	0.077	19:56
Air Blank	0.000	19:56
Subject Sample # 1	0.099	19:57
Breath Volume	3.152	19:57
Air Blank	0.000	19:58
Air Blank	0.000	19:59
Subject Sample # 2	0.085	20:00
Breath Volume	3.152	20:00
Air Blank	0.000	20:01
Control Test	0.077	20:01
Air Blank	0.000	20:02
Diagnostics Check	OK	20:02

Data Download Date/Time: 09/26/2018 07:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENLEY, WILLIAM, C
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: HENLEY, WILLIAM
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 08/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: MEDLEN

DATE/TIME OF TEST: 09/09/2018 01:17:58

Diagnostics Check	OK	01:11
Air Blank	0.000	01:11
Control Test	0.079	01:12
Air Blank	0.000	01:12
Subject Sample # 1	0.189	01:13
Breath Volume	2.570	01:13
Air Blank	0.000	01:13
Air Blank	0.000	01:15
Subject Sample # 2	0.191	01:16
Breath Volume	2.738	01:16
Air Blank	0.000	01:17
Control Test	0.078	01:17
Air Blank	0.000	01:17
Diagnostics Check	OK	01:17

Data Download Date/Time: 09/26/2018 07:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,AMANDA,L
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: DAVIS,AMANDA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 08/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: PAYNE

DATE/TIME OF TEST: 09/09/2018 05:02:41

Diagnostics Check	OK	04:57	
Air Blank	0.000	04:57	
Control Test	0.079	04:57	
Air Blank	0.000	04:58	
Subject Sample	# 1	05:01	Subject Test Refused
Breath Volume	0.000	05:01	
Air Blank	0.000	05:01	
Control Test	0.079	05:02	
Air Blank	0.000	05:02	
Diagnostics Check	OK	05:02	

Data Download Date/Time: 09/26/2018 07:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,AMANDA,L
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: DAVIS,AMANDA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 08/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: PAYNE

DATE/TIME OF TEST: 09/09/2018 05:15:43

Diagnostics Check	OK	05:08
Air Blank	0.000	05:09
Control Test	0.078	05:09
Air Blank	0.000	05:10
Subject Sample # 1	0.255	05:11
Breath Volume	3.578	05:11
Air Blank	0.000	05:11
Air Blank	0.000	05:13
Subject Sample # 2	0.253	05:14
Breath Volume	3.542	05:14
Air Blank	0.000	05:14
Control Test	0.077	05:15
Air Blank	0.000	05:15
Diagnostics Check	OK	05:15

Data Download Date/Time: 09/26/2018 07:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,MARCUS,E
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: ROEDEL,JUSTIN
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 08/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: MYERS

DATE/TIME OF TEST: 09/09/2018 17:57:15

Diagnostics Check	OK	17:50
Air Blank	0.000	17:51
Control Test	0.078	17:51
Air Blank	0.000	17:52
Subject Sample # 1	0.138	17:52
Breath Volume	2.253	17:52
Air Blank	0.000	17:53
Air Blank	0.000	17:55
Subject Sample # 2	0.133	17:55
Breath Volume	2.417	17:55
Air Blank	0.000	17:56
Control Test	0.078	17:56
Air Blank	0.000	17:57
Diagnostics Check	OK	17:57

Data Download Date/Time: 09/26/2018 07:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BYRD,BRANDON,C
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: BYRD,BRANDON
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 08/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 09/09/2018 21:04:52

Diagnostics Check	OK	20:58
Air Blank	0.000	20:58
Control Test	0.077	20:59
Air Blank	0.000	20:59
Subject Sample # 1	0.183	21:00
Breath Volume	4.093	21:00
Air Blank	0.000	21:00
Air Blank	0.000	21:02
Subject Sample # 2	0.185	21:03
Breath Volume	3.656	21:03
Air Blank	0.000	21:03
Control Test	0.077	21:04
Air Blank	0.000	21:04
Diagnostics Check	OK	21:04

Data Download Date/Time: 09/26/2018 07:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHORETTE,BRIAN,E
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: SHORETTE,BRIAN
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 08/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:28

SUBJECT'S LAST NAME: SILVA

DATE/TIME OF TEST: 09/09/2018 21:54:51

Diagnostics Check	OK	21:48
Air Blank	0.000	21:49
Control Test	0.077	21:49
Air Blank	0.000	21:49
Subject Sample # 1	0.157	21:50
Breath Volume	1.601	21:50
Air Blank	0.000	21:51
Air Blank	0.000	21:52
Subject Sample # 2	0.155	21:53
Breath Volume	1.585	21:53
Air Blank	0.000	21:53
Control Test	0.077	21:54
Air Blank	0.000	21:54
Diagnostics Check	OK	21:54

Data Download Date/Time: 09/26/2018 07:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BYRD,BRANDON,C
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: BYRD,BRANDON
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 08/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: HOLLINS

DATE/TIME OF TEST: 09/13/2018 01:36:33

Diagnostics Check	OK	01:29
Air Blank	0.000	01:30
Control Test	0.078	01:30
Air Blank	0.000	01:30
Subject Sample # 1	0.180	01:31
Breath Volume	1.992	01:31
Air Blank	0.000	01:32
Air Blank	0.000	01:34
Subject Sample # 2	0.185	01:35
Breath Volume	2.507	01:35
Air Blank	0.000	01:35
Control Test	0.077	01:35
Air Blank	0.000	01:36
Diagnostics Check	OK	01:36

Data Download Date/Time: 09/26/2018 07:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BYRD,BRANDON,C
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: BUZBEE,HEATHER
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 08/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: PARSONS

DATE/TIME OF TEST: 09/13/2018 01:56:13

Diagnostics Check	OK	01:49
Air Blank	0.000	01:49
Control Test	0.078	01:50
Air Blank	0.000	01:50
Subject Sample # 1	0.207	01:51
Breath Volume	3.718	01:51
Air Blank	0.000	01:52
Air Blank	0.000	01:53
Subject Sample # 2	0.211	01:54
Breath Volume	3.933	01:54
Air Blank	0.000	01:55
Control Test	0.077	01:55
Air Blank	0.000	01:56
Diagnostics Check	OK	01:56

Data Download Date/Time: 09/26/2018 07:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,HENDRICK,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MARTINEZ,HENDRICK
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: BRIDGES

DATE/TIME OF TEST: 09/14/2018 00:13:54

Diagnostics Check	OK	00:12	
Air Blank	0.000	00:13	
Control Test	0.072	00:13	Control Outside Tolerance
Air Blank	0.000	00:13	

Data Download Date/Time: 09/26/2018 07:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,HENDRICK,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MARTINEZ,HENDRICK
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: BRIDGES

DATE/TIME OF TEST: 09/14/2018 00:22:59

Diagnostics Check	OK	00:16
Air Blank	0.000	00:16
Control Test	0.078	00:17
Air Blank	0.000	00:17
Subject Sample # 1	0.173	00:18
Breath Volume	3.152	00:18
Air Blank	0.000	00:19
Air Blank	0.000	00:20
Subject Sample # 2	0.173	00:21
Breath Volume	2.859	00:21
Air Blank	0.000	00:22
Control Test	0.078	00:22
Air Blank	0.000	00:22
Diagnostics Check	OK	00:22

Data Download Date/Time: 09/26/2018 07:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENLEY, WILLIAM, C
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: HENLEY, WILLIAM
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 08/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: FELGER

DATE/TIME OF TEST: 09/16/2018 00:35:54

Diagnostics Check	OK	00:28
Air Blank	0.000	00:29
Control Test	0.078	00:29
Air Blank	0.000	00:30
Subject Sample # 1	0.151	00:31
Breath Volume	1.687	00:31
Air Blank	0.000	00:32
Air Blank	0.000	00:33
Subject Sample # 2	0.158	00:34
Breath Volume	1.589	00:34
Air Blank	0.000	00:35
Control Test	0.077	00:35
Air Blank	0.000	00:35
Diagnostics Check	OK	00:35

Data Download Date/Time: 09/26/2018 07:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELLEY,SAMUEL,L
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: SHELLEY,SAMUEL
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 08/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: TABAGO

DATE/TIME OF TEST: 09/16/2018 04:00:05

Diagnostics Check	OK	03:53
Air Blank	0.000	03:54
Control Test	0.079	03:54
Air Blank	0.000	03:55
Subject Sample # 1	0.166	03:55
Breath Volume	1.941	03:55
Air Blank	0.000	03:56
Air Blank	0.000	03:58
Subject Sample # 2	0.168	03:58
Breath Volume	2.402	03:58
Air Blank	0.000	03:59
Control Test	0.078	03:59
Air Blank	0.000	04:00
Diagnostics Check	OK	04:00

Data Download Date/Time: 09/26/2018 07:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUDSON,JOSHUA,K
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: HUDSON,JOSHUA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 08/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: ALLISON

DATE/TIME OF TEST: 09/16/2018 04:29:01

Diagnostics Check	OK	04:26	
Air Blank	0.000	04:26	
Control Test	0.079	04:26	
Air Blank	0.000	04:27	
Subject Sample	# 1	04:27	Subject Test Refused
Breath Volume	0.000	04:27	
Air Blank	0.000	04:28	
Control Test	0.078	04:28	
Air Blank	0.000	04:28	
Diagnostics Check	OK	04:29	

Data Download Date/Time: 09/26/2018 07:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BYRD,BRANDON,C
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: BYRD,BRANDON
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 08/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 09/17/2018 23:51:20

Diagnostics Check	OK	23:43
Air Blank	0.000	23:44
Control Test	0.079	23:44
Air Blank	0.000	23:45
Subject Sample # 1	0.038	23:46
Breath Volume	2.273	23:46
Air Blank	0.000	23:46
Air Blank	0.000	23:48
Subject Sample # 2	0.035	23:49
Breath Volume	2.765	23:49
Air Blank	0.000	23:50
Control Test	0.078	23:50
Air Blank	0.000	23:51
Diagnostics Check	OK	23:51

Data Download Date/Time: 09/26/2018 07:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,AMANDA,L
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: CALEY,TYLER
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 08/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 09/19/2018 02:59:28

Diagnostics Check	OK	02:52
Air Blank	0.000	02:53
Control Test	0.078	02:53
Air Blank	0.000	02:54
Subject Sample # 1	0.000	02:54
Breath Volume	1.625	02:54
Air Blank	0.000	02:55
Air Blank	0.000	02:57
Subject Sample # 2	0.000	02:58
Breath Volume	1.492	02:58
Air Blank	0.000	02:58
Control Test	0.079	02:58
Air Blank	0.000	02:59
Diagnostics Check	OK	02:59

Data Download Date/Time: 09/26/2018 07:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SALON,DENNIS,M
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: TOURNEY,KEITH
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 08/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 10:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:15

SUBJECT'S LAST NAME: CURRIE

DATE/TIME OF TEST: 09/19/2018 11:41:51

Diagnostics Check	OK	11:35
Air Blank	0.000	11:35
Control Test	0.077	11:36
Air Blank	0.000	11:36
Subject Sample # 1	0.262	11:37
Breath Volume	2.933	11:37
Air Blank	0.000	11:38
Air Blank	0.000	11:39
Subject Sample # 2	0.263	11:40
Breath Volume	1.984	11:40
Air Blank	0.000	11:40
Control Test	0.075	11:41
Air Blank	0.000	11:41
Diagnostics Check	OK	11:41

Data Download Date/Time: 09/26/2018 07:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARACLE, MATTHEW, D
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: COLLINS, CHASE
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 08/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:22

SUBJECT'S LAST NAME: HEARN

DATE/TIME OF TEST: 09/20/2018 17:07:00

Diagnostics Check	OK	16:58
Air Blank	0.000	16:59
Control Test	0.078	16:59
Air Blank	0.000	17:00
Subject Sample # 1	0.192	17:02
Breath Volume	1.531	17:02
Air Blank	0.000	17:03
Air Blank	0.000	17:04
Subject Sample # 2	0.197	17:05
Breath Volume	1.632	17:05
Air Blank	0.000	17:06
Control Test	0.077	17:06
Air Blank	0.000	17:06
Diagnostics Check	OK	17:07

Data Download Date/Time: 09/26/2018 07:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUDSON,JOSHUA,K
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: POSEY,BEN
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 08/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: HUDSPATH

DATE/TIME OF TEST: 09/21/2018 03:26:03

Diagnostics Check	OK	03:15
Air Blank	0.000	03:15
Control Test	0.078	03:16
Air Blank	0.000	03:16
Subject Sample # 1	0.055	03:19
Breath Volume	1.425	03:19
Air Blank	0.000	03:19
Air Blank	0.000	03:21
Subject Sample # 2	0.054	03:24
Breath Volume	2.160	03:24
Air Blank	0.000	03:25
Control Test	0.079	03:25
Air Blank	0.000	03:25
Diagnostics Check	OK	03:26

Data Download Date/Time: 09/26/2018 07:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOWLES, CONSTANCE, J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: KNOWLES, CONSTANCE
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: BRIGMAN

DATE/TIME OF TEST: 09/22/2018 01:22:12

Diagnostics Check	OK	01:15
Air Blank	0.000	01:15
Control Test	0.078	01:16
Air Blank	0.000	01:16
Subject Sample # 1	0.128	01:17
Breath Volume	1.687	01:17
Air Blank	0.000	01:18
Air Blank	0.000	01:20
Subject Sample # 2	0.125	01:20
Breath Volume	1.195	01:20
Air Blank	0.000	01:21
Control Test	0.078	01:21
Air Blank	0.000	01:22
Diagnostics Check	OK	01:22

Data Download Date/Time: 09/26/2018 07:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,AMANDA,L
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: DAVIS,AMANDA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 08/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 09/23/2018 19:36:35

Diagnostics Check	OK	19:33	
Air Blank	0.000	19:34	
Control Test	0.079	19:34	
Air Blank	0.000	19:35	
Subject Sample	# 1	19:35	Subject Test Refused
Breath Volume	0.000	19:35	
Air Blank	0.000	19:35	
Control Test	0.078	19:36	
Air Blank	0.000	19:36	
Diagnostics Check	OK	19:36	

Data Download Date/Time: 09/26/2018 07:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,AMANDA,L
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: LANDRY,KATHERINE
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 08/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: GOODALE

DATE/TIME OF TEST: 09/24/2018 00:39:59

Diagnostics Check	OK	00:33
Air Blank	0.000	00:33
Control Test	0.078	00:34
Air Blank	0.000	00:34
Subject Sample # 1	0.120	00:35
Breath Volume	2.078	00:35
Air Blank	0.000	00:35
Air Blank	0.000	00:37
Subject Sample # 2	0.120	00:38
Breath Volume	3.167	00:38
Air Blank	0.000	00:39
Control Test	0.078	00:39
Air Blank	0.000	00:39
Diagnostics Check	OK	00:39

Data Download Date/Time: 09/26/2018 07:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SECOR,BROC,A
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: SECOR,BROC
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 09/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: DAY

DATE/TIME OF TEST: 09/27/2018 02:55:42

Diagnostics Check	OK	02:47
Air Blank	0.000	02:47
Control Test	0.079	02:48
Air Blank	0.000	02:48
Subject Sample # 1	0.141	02:50
Breath Volume	2.097	02:50
Air Blank	0.000	02:50
Air Blank	0.000	02:52
Subject Sample # 2	0.136	02:54
Breath Volume	3.371	02:54
Air Blank	0.000	02:54
Control Test	0.080	02:55
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

Data Download Date/Time: 10/30/2018 07:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEETERS,OZZY,H
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: ISENBERG,KESLEY
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 14:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:13

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 09/27/2018 15:02:37

Diagnostics Check	OK	14:59	
Air Blank	0.000	14:59	
Control Test	0.079	15:00	
Air Blank	0.000	15:00	
Subject Sample	# 1	15:01	RFI Detect
Breath Volume	0.000	15:01	
Air Blank	0.000	15:01	
Control Test	0.079	15:02	
Air Blank	0.000	15:02	
Diagnostics Check	OK	15:02	

Data Download Date/Time: 10/30/2018 07:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEETERS,OZZY,H
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: ISENBERG,KESLEY
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 14:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:13

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 09/27/2018 15:07:27

Diagnostics Check	OK	15:04	
Air Blank	0.000	15:05	
Control Test	0.079	15:05	
Air Blank	0.000	15:05	
Subject Sample	# 1	15:06	Subject Test Refused
Breath Volume	0.000	15:06	
Air Blank	0.000	15:06	
Control Test	0.079	15:06	
Air Blank	0.000	15:07	
Diagnostics Check	OK	15:07	

Data Download Date/Time: 10/30/2018 07:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SECOR,BROC,A
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: SECOR,BROC
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 09/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: LINZY

DATE/TIME OF TEST: 09/28/2018 01:36:53

Diagnostics Check	OK	01:25	
Air Blank	0.000	01:25	
Control Test	0.079	01:26	
Air Blank	0.000	01:26	
Subject Sample	# 1 0.130	01:29	Volume Not Met
Breath Volume	0.281	01:29	
Air Blank	0.000	01:30	
Air Blank	0.000	01:32	
Subject Sample	# 2 0.135	01:35	Volume Not Met
Breath Volume	0.097	01:35	
Air Blank	0.000	01:35	
Control Test	0.077	01:36	
Air Blank	0.000	01:36	
Diagnostics Check	OK	01:36	

Data Download Date/Time: 10/30/2018 07:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUDSON,JOSHUA,K
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: POSEY,BEN
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: RICH

DATE/TIME OF TEST: 09/28/2018 20:58:18

Diagnostics Check	OK	20:52
Air Blank	0.000	20:52
Control Test	0.079	20:52
Air Blank	0.000	20:53
Subject Sample # 1	0.214	20:53
Breath Volume	1.101	20:53
Air Blank	0.000	20:54
Air Blank	0.000	20:56
Subject Sample # 2	0.203	20:56
Breath Volume	1.648	20:56
Air Blank	0.000	20:57
Control Test	0.078	20:57
Air Blank	0.000	20:58
Diagnostics Check	OK	20:58

Data Download Date/Time: 10/30/2018 07:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUDSON,JOSHUA,K
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: POSEY,BEN
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: HULETT

DATE/TIME OF TEST: 09/29/2018 03:22:45

Diagnostics Check	OK	03:16
Air Blank	0.000	03:16
Control Test	0.080	03:16
Air Blank	0.000	03:17
Subject Sample # 1	0.185	03:18
Breath Volume	2.363	03:18
Air Blank	0.000	03:18
Air Blank	0.000	03:20
Subject Sample # 2	0.184	03:21
Breath Volume	2.582	03:21
Air Blank	0.000	03:21
Control Test	0.079	03:22
Air Blank	0.000	03:22
Diagnostics Check	OK	03:22

Data Download Date/Time: 10/30/2018 07:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUDSON,JOSHUA,K
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: POSEY,BEN
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: PLANT

DATE/TIME OF TEST: 09/30/2018 05:13:59

Diagnostics Check	OK	05:07
Air Blank	0.000	05:07
Control Test	0.079	05:08
Air Blank	0.000	05:08
Subject Sample # 1	0.068	05:09
Breath Volume	2.394	05:09
Air Blank	0.000	05:09
Air Blank	0.000	05:11
Subject Sample # 2	0.065	05:12
Breath Volume	1.910	05:12
Air Blank	0.000	05:13
Control Test	0.080	05:13
Air Blank	0.000	05:13
Diagnostics Check	OK	05:13

Data Download Date/Time: 10/30/2018 07:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WENDT,HENDRIK,J
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: HUGHES,DAVID
ARRESTING OFFICER AGENCY: Alachua PD
LAST AGENCY INSPECTION: 08/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: ROCHELLE

DATE/TIME OF TEST: 09/09/2018 02:48:52

Diagnostics Check	OK	02:37	
Air Blank	0.000	02:37	
Control Test	0.078	02:38	
Air Blank	0.000	02:38	
Subject Sample	# 1	02:41	No Sample Provided
Breath Volume	0.000	02:41	
Air Blank	0.000	02:42	
Air Blank	0.000	02:44	
Subject Sample	# 2	02:47	No Sample Provided
Breath Volume	0.000	02:47	
Air Blank	0.000	02:47	
Control Test	0.080	02:48	
Air Blank	0.000	02:48	
Diagnostics Check	OK	02:48	

Data Download Date/Time: 09/24/2018 09:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SABO,JESSICA,L
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SABO,JESSICA
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: ROJANO FERNANDEZ

DATE/TIME OF TEST: 09/11/2018 04:24:25

Diagnostics Check	OK	04:18
Air Blank	0.000	04:18
Control Test	0.079	04:19
Air Blank	0.000	04:19
Subject Sample # 1	0.128	04:20
Breath Volume	1.605	04:20
Air Blank	0.000	04:20
Air Blank	0.000	04:22
Subject Sample # 2	0.128	04:22
Breath Volume	1.949	04:22
Air Blank	0.000	04:23
Control Test	0.080	04:23
Air Blank	0.000	04:24
Diagnostics Check	OK	04:24

Data Download Date/Time: 09/24/2018 09:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BREWINGTON, RAMON, C
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BREWINGTON, RAMON
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 06:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:20

SUBJECT'S LAST NAME: MULLINS

DATE/TIME OF TEST: 09/11/2018 09:01:42

Diagnostics Check	OK	08:50	
Air Blank	0.000	08:50	
Control Test	0.079	08:51	
Air Blank	0.000	08:51	
Subject Sample	# 1	08:54	No Sample Provided
Breath Volume	0.000	08:54	
Air Blank	0.000	08:55	
Air Blank	0.000	08:57	
Subject Sample	# 2	09:00	No Sample Provided
Breath Volume	0.000	09:00	
Air Blank	0.000	09:00	
Control Test	0.080	09:01	
Air Blank	0.000	09:01	
Diagnostics Check	OK	09:01	

Data Download Date/Time: 09/24/2018 09:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNS, WILLIAM, A
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: MCCANDLESS, MATTHEW
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 08/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 09/14/2018 04:41:36

Diagnostics Check	OK	04:29	
Air Blank	0.000	04:30	
Control Test	0.079	04:30	
Air Blank	0.000	04:30	
Subject Sample	# 1 0.209	04:34	Volume Not Met
Breath Volume	0.300	04:34	
Air Blank	0.000	04:35	
Air Blank	0.000	04:36	
Subject Sample	# 2 0.209	04:39	Volume Not Met
Breath Volume	0.148	04:39	
Air Blank	0.000	04:40	
Control Test	0.074	04:40	Control Outside Tolerance
Air Blank	0.000	04:41	

Data Download Date/Time: 09/24/2018 09:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNS,WILLIAM,A
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: MCCANDLESS,MATTHEW
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 08/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 09/14/2018 04:45:49

Diagnostics Check	OK	04:44	
Air Blank	0.000	04:44	
Control Test	0.074	04:45	Control Outside Tolerance
Air Blank	0.000	04:45	

Data Download Date/Time: 09/24/2018 09:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNS, WILLIAM, A
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: MCCANDLESS, MATTHEW
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 08/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 09/14/2018 04:54:49

Diagnostics Check	OK	04:51	
Air Blank	0.000	04:52	
Control Test	0.078	04:52	
Air Blank	0.000	04:53	
Subject Sample	# 1	04:53	Subject Test Refused
Breath Volume	0.000	04:53	
Air Blank	0.000	04:53	
Control Test	0.080	04:54	
Air Blank	0.000	04:54	
Diagnostics Check	OK	04:54	

Data Download Date/Time: 09/24/2018 09:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELCH,CHAD,J
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GUYTON,JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 08/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: DURRANCE

DATE/TIME OF TEST: 09/20/2018 23:37:20

Diagnostics Check	OK	23:25	
Air Blank	0.000	23:26	
Control Test	0.079	23:26	
Air Blank	0.000	23:27	
Subject Sample	# 1	23:30	No Sample Provided
Breath Volume	0.000	23:30	
Air Blank	0.000	23:30	
Air Blank	0.000	23:32	
Subject Sample	# 2	23:35	No Sample Provided
Breath Volume	0.000	23:35	
Air Blank	0.000	23:36	
Control Test	0.080	23:36	
Air Blank	0.000	23:37	
Diagnostics Check	OK	23:37	

Data Download Date/Time: 09/24/2018 09:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELCH,CHAD,J
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GUYTON,JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 08/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: BURNHAM

DATE/TIME OF TEST: 09/21/2018 04:37:57

Diagnostics Check	OK	04:26	
Air Blank	0.000	04:26	
Control Test	0.078	04:27	
Air Blank	0.000	04:27	
Subject Sample	# 1	04:30	No Sample Provided
Breath Volume	0.000	04:30	
Air Blank	0.000	04:31	
Air Blank	0.000	04:33	
Subject Sample	# 2	04:36	No Sample Provided
Breath Volume	0.000	04:36	
Air Blank	0.000	04:36	
Control Test	0.080	04:37	
Air Blank	0.000	04:37	
Diagnostics Check	OK	04:37	

Data Download Date/Time: 09/24/2018 09:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GATES,DAVID,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: PHELAN,MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 07:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:40

SUBJECT'S LAST NAME: BERRIOS

DATE/TIME OF TEST: 09/22/2018 09:24:27

Diagnostics Check	OK	09:11	
Air Blank	0.000	09:11	
Control Test	0.079	09:11	
Air Blank	0.000	09:12	
Subject Sample	# 1 0.118	09:15	Volume Not Met
Breath Volume	1.066	09:15	
Air Blank	0.000	09:16	
Air Blank	0.000	09:17	
Subject Sample	# 2 0.118	09:19	
Breath Volume	1.714	09:19	
Air Blank	0.000	09:20	
Air Blank	0.000	09:22	
Subject Sample	# 3 0.120	09:22	
Breath Volume	1.785	09:22	
Air Blank	0.000	09:23	
Control Test	0.078	09:23	
Air Blank	0.000	09:24	
Diagnostics Check	OK	09:24	

Data Download Date/Time: 09/24/2018 09:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WENDT,HENDRIK,J
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: FUENTES,MELISSA
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 09/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: AVERY-HOISINGTON

DATE/TIME OF TEST: 09/28/2018 02:46:18

Diagnostics Check	OK	02:39
Air Blank	0.000	02:39
Control Test	0.078	02:39
Air Blank	0.000	02:40
Subject Sample # 1	0.182	02:40
Breath Volume	2.273	02:40
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.176	02:44
Breath Volume	2.566	02:44
Air Blank	0.000	02:45
Control Test	0.079	02:45
Air Blank	0.000	02:46
Diagnostics Check	OK	02:46

Data Download Date/Time: 10/26/2018 11:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORGAN,DARRAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MORGAN,DARRAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 14:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:09

SUBJECT'S LAST NAME: CISNEROS

DATE/TIME OF TEST: 09/28/2018 15:48:43

Diagnostics Check	OK	15:42
Air Blank	0.000	15:42
Control Test	0.079	15:42
Air Blank	0.000	15:43
Subject Sample # 1	0.191	15:44
Breath Volume	2.894	15:44
Air Blank	0.000	15:44
Air Blank	0.000	15:46
Subject Sample # 2	0.173	15:47
Breath Volume	1.257	15:47
Air Blank	0.000	15:47
Control Test	0.079	15:48
Air Blank	0.000	15:48
Diagnostics Check	OK	15:48

Data Download Date/Time: 10/26/2018 11:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRINSKO,THOMAS,K
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: ADKINS,MICHAEL
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 09/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 09/30/2018 03:44:21

Diagnostics Check	OK	03:41	
Air Blank	0.000	03:41	
Control Test	0.079	03:42	
Air Blank	0.000	03:42	
Subject Sample	# 1	03:42	Subject Test Refused
Breath Volume	0.000	03:42	
Air Blank	0.000	03:43	
Control Test	0.079	03:43	
Air Blank	0.000	03:44	
Diagnostics Check	OK	03:44	

Data Download Date/Time: 10/26/2018 11:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRINSKO,THOMAS,K
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GUYTON,JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 09/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: FLEENOR

DATE/TIME OF TEST: 09/30/2018 04:23:09

Diagnostics Check	OK	04:16
Air Blank	0.000	04:17
Control Test	0.079	04:17
Air Blank	0.000	04:18
Subject Sample # 1	0.194	04:18
Breath Volume	1.710	04:18
Air Blank	0.000	04:19
Air Blank	0.000	04:21
Subject Sample # 2	0.186	04:21
Breath Volume	1.644	04:21
Air Blank	0.000	04:22
Control Test	0.079	04:22
Air Blank	0.000	04:23
Diagnostics Check	OK	04:23

Data Download Date/Time: 10/26/2018 11:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYNN,BAMBI,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SPENCER,RAYMOND
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG715202
DRY GAS EXPIRY DATE: 06/01/2019
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 09/01/2018 23:30:50

Diagnostics Check	OK	23:26	
Air Blank	0.000	23:26	
Control Test	0.077	23:26	
Air Blank	0.000	23:27	
Subject Sample	# 1	23:29	Subject Test Refused
Breath Volume	0.000	23:29	
Air Blank	0.000	23:29	
Control Test	0.078	23:30	
Air Blank	0.000	23:30	
Diagnostics Check	OK	23:30	

Data Download Date/Time: 09/22/2018 00:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,COREY,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CALABRO,ALFRED
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG715202
DRY GAS EXPIRY DATE: 06/01/2019
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: SNOW

DATE/TIME OF TEST: 09/02/2018 02:20:38

Diagnostics Check	OK	02:12	
Air Blank	0.000	02:12	
Control Test	0.077	02:12	
Air Blank	0.000	02:13	
Subject Sample	# 1	02:16	No Sample Provided
Breath Volume	0.000	02:16	
Air Blank	0.000	02:17	
Air Blank	0.000	02:19	
Subject Sample	# 2	02:19	Subject Test Refused
Breath Volume	0.000	02:19	
Air Blank	0.000	02:19	
Control Test	0.078	02:20	
Air Blank	0.000	02:20	
Diagnostics Check	OK	02:20	

Data Download Date/Time: 09/22/2018 00:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,COREY,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: BERRY,COREY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 09/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG715202
DRY GAS EXPIRY DATE: 06/01/2019
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: MARRAFINO

DATE/TIME OF TEST: 09/23/2018 00:58:17

Diagnostics Check	OK	00:50
Air Blank	0.000	00:51
Control Test	0.082	00:51
Air Blank	0.000	00:52
Subject Sample # 1	0.178	00:53
Breath Volume	1.972	00:53
Air Blank	0.000	00:54
Air Blank	0.000	00:56
Subject Sample # 2	0.178	00:56
Breath Volume	2.253	00:56
Air Blank	0.000	00:57
Control Test	0.081	00:57
Air Blank	0.000	00:58
Diagnostics Check	OK	00:58

Data Download Date/Time: 11/03/2018 00:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMIREZ,MICHAEL,E
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: THOMPSON,WILLIAM
ARRESTING OFFICER AGENCY: Milton PD
LAST AGENCY INSPECTION: 09/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG715202
DRY GAS EXPIRY DATE: 06/01/2019
VIOLATION CODE: DUI
ARREST TIME: 18:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: COON

DATE/TIME OF TEST: 09/24/2018 19:37:37

Diagnostics Check	OK	19:37	
Air Blank		19:37	RFI Detect
Air Blank		19:37	RFI Detect

Data Download Date/Time: 11/03/2018 00:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMIREZ,MICHAEL,E
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: THOMPSON,WILLIAM
ARRESTING OFFICER AGENCY: Milton PD
LAST AGENCY INSPECTION: 09/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG715202
DRY GAS EXPIRY DATE: 06/01/2019
VIOLATION CODE: DUI
ARREST TIME: 18:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: COON

DATE/TIME OF TEST: 09/24/2018 19:45:38

Diagnostics Check	OK	19:39	
Air Blank	0.000	19:40	
Control Test	0.083	19:40	
Air Blank	0.000	19:41	
Subject Sample # 1	0.000	19:41	
Breath Volume	2.687	19:41	
Air Blank	0.000	19:42	
Air Blank		19:44	RFI Detect
Air Blank	0.000	19:44	
Control Test	0.083	19:45	
Air Blank	0.000	19:45	
Diagnostics Check	OK	19:45	

Data Download Date/Time: 11/03/2018 00:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWELL,DAVID,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ROWELL,DAVID
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 09/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG715202
DRY GAS EXPIRY DATE: 06/01/2019
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: FERRARI CORDERO

DATE/TIME OF TEST: 09/28/2018 00:57:38

Diagnostics Check	OK	00:50
Air Blank	0.000	00:51
Control Test	0.082	00:51
Air Blank	0.000	00:52
Subject Sample # 1	0.180	00:53
Breath Volume	3.398	00:53
Air Blank	0.000	00:53
Air Blank	0.000	00:55
Subject Sample # 2	0.173	00:56
Breath Volume	3.550	00:56
Air Blank	0.000	00:56
Control Test	0.082	00:57
Air Blank	0.000	00:57
Diagnostics Check	OK	00:57

Data Download Date/Time: 11/03/2018 00:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,BENJAMIN,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WARD,BENJAMIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG715202
DRY GAS EXPIRY DATE: 06/01/2019
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: SANFORD

DATE/TIME OF TEST: 09/28/2018 05:01:37

Diagnostics Check	OK	04:55
Air Blank	0.000	04:56
Control Test	0.082	04:56
Air Blank	0.000	04:57
Subject Sample # 1	0.000	04:57
Breath Volume	1.539	04:57
Air Blank	0.000	04:57
Air Blank	0.000	04:59
Subject Sample # 2	0.000	05:00
Breath Volume	1.429	05:00
Air Blank	0.000	05:00
Control Test	0.081	05:01
Air Blank	0.000	05:01
Diagnostics Check	OK	05:01

Data Download Date/Time: 11/03/2018 00:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERHUNE,JEREMY,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: TERHUNE,JEREMY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 09/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG715202
DRY GAS EXPIRY DATE: 06/01/2019
VIOLATION CODE: DUI
ARREST TIME: 22:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 09/29/2018 00:13:59

Diagnostics Check	OK	00:07
Air Blank	0.000	00:08
Control Test	0.082	00:08
Air Blank	0.000	00:08
Subject Sample # 1	0.037	00:09
Breath Volume	2.132	00:09
Air Blank	0.000	00:10
Air Blank	0.000	00:11
Subject Sample # 2	0.033	00:12
Breath Volume	1.476	00:12
Air Blank	0.000	00:13
Control Test	0.081	00:13
Air Blank	0.000	00:13
Diagnostics Check	OK	00:13

Data Download Date/Time: 11/03/2018 00:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAINE,JACK,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BLAINE,JACK
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG715202
DRY GAS EXPIRY DATE: 06/01/2019
VIOLATION CODE: DUI
ARREST TIME: 18:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:12

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 09/29/2018 18:51:58

Diagnostics Check	OK	18:48	
Air Blank	0.000	18:48	
Control Test	0.082	18:49	
Air Blank	0.000	18:49	
Subject Sample	# 1	18:50	Subject Test Refused
Breath Volume	0.000	18:50	
Air Blank	0.000	18:51	
Control Test	0.082	18:51	
Air Blank	0.000	18:51	
Diagnostics Check	OK	18:51	

Data Download Date/Time: 11/03/2018 00:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOYE,JAMES,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MOYE,JAMES
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 09/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG715202
DRY GAS EXPIRY DATE: 06/01/2019
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: ROSS

DATE/TIME OF TEST: 09/30/2018 00:37:59

Diagnostics Check	OK	00:35	
Air Blank	0.000	00:35	
Control Test	0.081	00:35	
Air Blank	0.000	00:36	
Subject Sample	# 1	00:36	Subject Test Refused
Breath Volume	0.000	00:36	
Air Blank	0.000	00:37	
Control Test	0.082	00:37	
Air Blank	0.000	00:37	
Diagnostics Check	OK	00:37	

Data Download Date/Time: 11/03/2018 00:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,DANIEL,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WAGNER,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: BOWEN

DATE/TIME OF TEST: 09/01/2018 02:12:35

Diagnostics Check	OK	02:00	
Air Blank	0.000	02:00	
Control Test	0.081	02:01	
Air Blank	0.000	02:01	
Subject Sample	# 1 0.164	02:04	Volume Not Met
Breath Volume	0.828	02:04	
Air Blank	0.000	02:05	
Air Blank	0.000	02:07	
Subject Sample	# 2 0.162	02:07	
Breath Volume	1.855	02:07	
Air Blank	0.000	02:08	
Air Blank	0.000	02:10	
Subject Sample	# 3 0.173	02:10	
Breath Volume	1.691	02:10	
Air Blank	0.000	02:11	
Control Test	0.081	02:11	
Air Blank	0.000	02:12	
Diagnostics Check	OK	02:12	

Data Download Date/Time: 09/27/2018 06:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CREWS,MICHAEL,B
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: HARVEY,DAVID
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 08/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: SAPP-EDWARDS

DATE/TIME OF TEST: 09/02/2018 03:17:12

Diagnostics Check	OK	03:10
Air Blank	0.000	03:11
Control Test	0.081	03:11
Air Blank	0.000	03:11
Subject Sample # 1	0.065	03:12
Breath Volume	2.539	03:12
Air Blank	0.000	03:13
Air Blank	0.000	03:15
Subject Sample # 2	0.065	03:15
Breath Volume	2.488	03:15
Air Blank	0.000	03:16
Control Test	0.081	03:16
Air Blank	0.000	03:17
Diagnostics Check	OK	03:17

Data Download Date/Time: 09/27/2018 06:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CREWS,MICHAEL,B
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: WYNNE,AUSTIN
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 08/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 09/02/2018 03:25:19

Diagnostics Check	OK	03:19
Air Blank	0.000	03:19
Control Test	0.081	03:19
Air Blank	0.000	03:20
Subject Sample # 1	0.025	03:20
Breath Volume	1.500	03:20
Air Blank	0.000	03:21
Air Blank	0.000	03:23
Subject Sample # 2	0.024	03:23
Breath Volume	1.269	03:23
Air Blank	0.000	03:24
Control Test	0.081	03:24
Air Blank	0.000	03:25
Diagnostics Check	OK	03:25

Data Download Date/Time: 09/27/2018 06:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAURES,BRANDYN,K
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BAURES,BRANDYN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 20:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: LATESSA

DATE/TIME OF TEST: 09/03/2018 21:26:35

Diagnostics Check	OK	21:20
Air Blank	0.000	21:21
Control Test	0.081	21:21
Air Blank	0.000	21:21
Subject Sample # 1	0.000	21:22
Breath Volume	2.832	21:22
Air Blank	0.000	21:22
Air Blank	0.000	21:24
Subject Sample # 2	0.000	21:25
Breath Volume	2.847	21:25
Air Blank	0.000	21:25
Control Test	0.081	21:25
Air Blank	0.000	21:26
Diagnostics Check	OK	21:26

Data Download Date/Time: 09/27/2018 06:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CREWS,MICHAEL,B
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: WORTHINGTON,JOSHUA
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 08/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: TROUTNER

DATE/TIME OF TEST: 09/06/2018 00:56:31

Diagnostics Check	OK	00:50
Air Blank	0.000	00:50
Control Test	0.082	00:51
Air Blank	0.000	00:51
Subject Sample # 1	0.102	00:52
Breath Volume	2.355	00:52
Air Blank	0.000	00:52
Air Blank	0.000	00:54
Subject Sample # 2	0.102	00:54
Breath Volume	2.445	00:54
Air Blank	0.000	00:55
Control Test	0.082	00:55
Air Blank	0.000	00:56
Diagnostics Check	OK	00:56

Data Download Date/Time: 09/27/2018 06:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CREWS,MICHAEL,B
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: CREWS,MICHAEL
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 08/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: HERMESMAN

DATE/TIME OF TEST: 09/07/2018 02:35:45

Diagnostics Check	OK	02:29
Air Blank	0.000	02:29
Control Test	0.082	02:30
Air Blank	0.000	02:30
Subject Sample # 1	0.114	02:31
Breath Volume	1.996	02:31
Air Blank	0.000	02:32
Air Blank	0.000	02:33
Subject Sample # 2	0.105	02:34
Breath Volume	2.128	02:34
Air Blank	0.000	02:34
Control Test	0.082	02:35
Air Blank	0.000	02:35
Diagnostics Check	OK	02:35

Data Download Date/Time: 09/27/2018 06:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATES,BRIAN,C
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BATES,BRIAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 09/14/2018 00:51:53

Diagnostics Check	OK	00:44
Air Blank	0.000	00:45
Control Test	0.081	00:45
Air Blank	0.000	00:46
Subject Sample # 1	0.000	00:46
Breath Volume	1.644	00:46
Air Blank	0.000	00:47
Air Blank	0.000	00:49
Subject Sample # 2	0.000	00:50
Breath Volume	1.578	00:50
Air Blank	0.000	00:50
Control Test	0.079	00:51
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

Data Download Date/Time: 09/27/2018 06:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUFORD,DAVID,A
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: HARVEY,DAVID
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 08/29/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 10:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:23

SUBJECT'S LAST NAME: DUFORD

DATE/TIME OF TEST: 09/15/2018 11:50:06

Diagnostics Check	OK	11:43
Air Blank	0.000	11:43
Control Test	0.082	11:44
Air Blank	0.000	11:44
Subject Sample # 1	0.042	11:45
Breath Volume	1.179	11:45
Air Blank	0.000	11:46
Air Blank	0.000	11:47
Subject Sample # 2	0.040	11:48
Breath Volume	2.003	11:48
Air Blank	0.000	11:49
Control Test	0.081	11:49
Air Blank	0.000	11:50
Diagnostics Check	OK	11:50

Data Download Date/Time: 09/27/2018 06:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CREWS,MICHAEL,B
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: CREWS,MICHAEL
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 08/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 08:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:50

SUBJECT'S LAST NAME: THOMAS JR

DATE/TIME OF TEST: 09/19/2018 09:17:38

Diagnostics Check	OK	09:10
Air Blank	0.000	09:11
Control Test	0.081	09:11
Air Blank	0.000	09:11
Subject Sample # 1	0.000	09:12
Breath Volume	2.246	09:12
Air Blank	0.000	09:12
Air Blank	0.000	09:14
Subject Sample # 2	0.000	09:16
Breath Volume	1.902	09:16
Air Blank	0.000	09:16
Control Test	0.082	09:17
Air Blank	0.000	09:17
Diagnostics Check	OK	09:17

Data Download Date/Time: 09/27/2018 06:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUSGROVE,DUSTIN,K
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MUSGROVE,DUSTIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 09:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:18

SUBJECT'S LAST NAME: PRIVETTE

DATE/TIME OF TEST: 09/19/2018 09:46:34

Diagnostics Check	OK	09:40
Air Blank	0.000	09:40
Control Test	0.081	09:41
Air Blank	0.000	09:41
Subject Sample # 1	0.000	09:42
Breath Volume	1.199	09:42
Air Blank	0.000	09:42
Air Blank	0.000	09:44
Subject Sample # 2	0.000	09:45
Breath Volume	1.371	09:45
Air Blank	0.000	09:45
Control Test	0.078	09:45
Air Blank	0.000	09:46
Diagnostics Check	OK	09:46

Data Download Date/Time: 09/27/2018 06:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000779

INSTRUMENT REGISTERED TO: MADISON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LLANES,GABRIEL,M
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: SANDERS,ROBERT
ARRESTING OFFICER AGENCY: Madison PD
LAST AGENCY INSPECTION: 08/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A3
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:53

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 09/02/2018 21:27:47

Diagnostics Check	OK	21:21
Air Blank	0.000	21:21
Control Test	0.080	21:21
Air Blank	0.000	21:22
Subject Sample # 1	0.158	21:23
Breath Volume	2.515	21:23
Air Blank	0.000	21:23
Air Blank	0.000	21:25
Subject Sample # 2	0.155	21:26
Breath Volume	2.644	21:26
Air Blank	0.000	21:26
Control Test	0.078	21:27
Air Blank	0.000	21:27
Diagnostics Check	OK	21:27

Data Download Date/Time: 09/27/2018 10:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000779

INSTRUMENT REGISTERED TO: MADISON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LLANES,GABRIEL,M
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: VOLLERTSEN,CRAIG
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 08/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A3
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 10:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: MORTON

DATE/TIME OF TEST: 09/02/2018 23:09:36

Diagnostics Check	OK	23:03
Air Blank	0.000	23:03
Control Test	0.079	23:04
Air Blank	0.000	23:04
Subject Sample # 1	0.140	23:05
Breath Volume	2.656	23:05
Air Blank	0.000	23:05
Air Blank	0.000	23:07
Subject Sample # 2	0.132	23:08
Breath Volume	1.277	23:08
Air Blank	0.000	23:08
Control Test	0.078	23:08
Air Blank	0.000	23:09
Diagnostics Check	OK	23:09

Data Download Date/Time: 09/27/2018 10:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000779

INSTRUMENT REGISTERED TO: MADISON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOREY,FREDERICK,J
BREATH TEST OPERATOR AGENCY: Madison County SO
ARRESTING OFFICER: PIKE,BENJAMIN
ARRESTING OFFICER AGENCY: Madison County SO
LAST AGENCY INSPECTION: 08/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A3
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 06:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:40

SUBJECT'S LAST NAME: FREEBURN

DATE/TIME OF TEST: 09/18/2018 07:11:29

Diagnostics Check	OK	07:04
Air Blank	0.000	07:05
Control Test	0.080	07:05
Air Blank	0.000	07:05
Subject Sample # 1	0.135	07:06
Breath Volume	3.718	07:06
Air Blank	0.000	07:07
Air Blank	0.000	07:09
Subject Sample # 2	0.138	07:09
Breath Volume	3.769	07:09
Air Blank	0.000	07:10
Control Test	0.077	07:10
Air Blank	0.000	07:11
Diagnostics Check	OK	07:11

Data Download Date/Time: 09/27/2018 10:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUSGROVE,DUSTIN,K
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MUSGROVE,DUSTIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: BRANTLEY

DATE/TIME OF TEST: 09/01/2018 03:09:54

Diagnostics Check	OK	03:00
Air Blank	0.000	03:01
Control Test	0.083	03:01
Air Blank	0.000	03:02
Subject Sample # 1	0.153	03:03
Breath Volume	1.554	03:03
Air Blank	0.000	03:04
Air Blank	0.000	03:06
Subject Sample # 2	0.133	03:07
Breath Volume	2.691	03:07
Air Blank	0.000	03:08
Control Test	0.082	03:09
Air Blank	0.000	03:09
Diagnostics Check	OK	03:09

Data Download Date/Time: 11/08/2018 11:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUSGROVE,DUSTIN,K
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MUSGROVE,DUSTIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: ANDREOLI

DATE/TIME OF TEST: 09/01/2018 23:38:12

Diagnostics Check	OK	23:29
Air Blank	0.000	23:29
Control Test	0.081	23:30
Air Blank	0.000	23:30
Subject Sample # 1	0.128	23:31
Breath Volume	1.542	23:31
Air Blank	0.000	23:32
Air Blank	0.000	23:34
Subject Sample # 2	0.126	23:36
Breath Volume	1.824	23:36
Air Blank	0.000	23:37
Control Test	0.077	23:37
Air Blank	0.000	23:38
Diagnostics Check	OK	23:38

Data Download Date/Time: 11/08/2018 11:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SINGLETARY,MATHEW,C
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SINGLETARY,MATHEW
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: WHEELER

DATE/TIME OF TEST: 09/02/2018 01:59:43

Diagnostics Check	OK	01:58	
Air Blank	0.000	01:58	
Control Test	0.083	01:59	
Air Blank		01:59	RFI Detect
Air Blank		01:59	RFI Detect

Data Download Date/Time: 11/08/2018 11:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SINGLETARY, MATHEW, C
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SINGLETARY, MATHEW
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: WHEELER

DATE/TIME OF TEST: 09/02/2018 02:05:23

Diagnostics Check	OK	02:02	
Air Blank	0.000	02:03	
Control Test	0.082	02:03	
Air Blank	0.000	02:03	
Subject Sample	# 1	02:04	Subject Test Refused
Breath Volume	0.000	02:04	
Air Blank	0.000	02:04	
Control Test	0.083	02:04	
Air Blank	0.000	02:05	
Diagnostics Check	OK	02:05	

Data Download Date/Time: 11/08/2018 11:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FALGOUT, DAVID, G
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: CRAIG, S
ARRESTING OFFICER AGENCY: Live Oak PD
LAST AGENCY INSPECTION: 08/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: STOVER

DATE/TIME OF TEST: 09/05/2018 02:11:24

Diagnostics Check	OK	02:01
Air Blank	0.000	02:01
Control Test	0.082	02:01
Air Blank	0.000	02:02
Subject Sample # 1	0.177	02:03
Breath Volume	1.816	02:03
Air Blank	0.000	02:04
Air Blank	0.000	02:06
Subject Sample # 2	0.155	02:06
Breath Volume	1.460	02:06
Air Blank	0.000	02:07
Air Blank	0.000	02:09
Subject Sample # 3	0.150	02:09
Breath Volume	1.328	02:09
Air Blank	0.000	02:10
Control Test	0.081	02:10
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

Data Download Date/Time: 11/08/2018 11:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUGHES,ALFRED,L
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: CRAIG,SHAWN
ARRESTING OFFICER AGENCY: Live Oak PD
LAST AGENCY INSPECTION: 08/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: HERNANDEZ MARTINEZ

DATE/TIME OF TEST: 09/16/2018 03:54:11

Diagnostics Check	OK	03:49	
Air Blank	0.000	03:49	
Control Test	0.083	03:50	
Air Blank	0.000	03:50	
Subject Sample # 1	0.225	03:52	
Breath Volume	1.800	03:52	
Air Blank		03:53	Purge Fail
Air Blank	0.000	03:54	

Data Download Date/Time: 11/08/2018 11:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUGHES,ALFRED,L
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: CRAIG,SHAWN
ARRESTING OFFICER AGENCY: Live Oak PD
LAST AGENCY INSPECTION: 08/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: HERNANDEZ MARTINEZ

DATE/TIME OF TEST: 09/16/2018 04:03:13

Diagnostics Check	OK	03:56
Air Blank	0.000	03:56
Control Test	0.079	03:57
Air Blank	0.000	03:57
Subject Sample # 1	0.221	03:58
Breath Volume	2.105	03:58
Air Blank	0.000	03:59
Air Blank	0.000	04:00
Subject Sample # 2	0.220	04:01
Breath Volume	2.117	04:01
Air Blank	0.000	04:02
Control Test	0.080	04:02
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time: 11/08/2018 11:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FALGOUT, DAVID, G
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: CHAUNCEY, R
ARRESTING OFFICER AGENCY: Live Oak PD
LAST AGENCY INSPECTION: 08/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:36

SUBJECT'S LAST NAME: RAULERSON

DATE/TIME OF TEST: 09/17/2018 21:02:57

Diagnostics Check	OK	20:58	
Air Blank	0.000	20:58	
Control Test	0.082	20:59	
Air Blank	0.000	20:59	
Subject Sample	# 1	21:01	Subject Test Refused
Breath Volume	1.218	21:01	
Air Blank	0.000	21:02	
Control Test	0.083	21:02	
Air Blank	0.000	21:02	
Diagnostics Check	OK	21:02	

Data Download Date/Time: 11/08/2018 11:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATES,BRIAN,C
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BATES,BRIAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:51

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 09/26/2018 22:18:42

Diagnostics Check	OK	22:11
Air Blank	0.000	22:11
Control Test	0.083	22:12
Air Blank	0.000	22:12
Subject Sample # 1	0.190	22:13
Breath Volume	2.023	22:13
Air Blank	0.000	22:14
Air Blank	0.000	22:16
Subject Sample # 2	0.183	22:16
Breath Volume	1.578	22:16
Air Blank	0.000	22:17
Control Test	0.080	22:18
Air Blank	0.000	22:18
Diagnostics Check	OK	22:18

Data Download Date/Time: 11/08/2018 11:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIEGEL,TIMOTHY,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: DURR,KYLE
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 09/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: RETHMAN

DATE/TIME OF TEST: 09/29/2018 00:01:37

Diagnostics Check	OK	23:55
Air Blank	0.000	23:55
Control Test	0.082	23:55
Air Blank	0.000	23:56
Subject Sample # 1	0.180	23:56
Breath Volume	3.351	23:56
Air Blank	0.000	23:57
Air Blank	0.000	23:59
Subject Sample # 2	0.177	23:59
Breath Volume	1.863	23:59
Air Blank	0.000	00:00
Control Test	0.080	00:01
Air Blank	0.000	00:01
Diagnostics Check	OK	00:01

Data Download Date/Time: 11/08/2018 11:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIEGEL,TIMOTHY,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: RIEGEL,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: DE LEON US

DATE/TIME OF TEST: 09/30/2018 04:41:17

Diagnostics Check	OK	04:38	
Air Blank	0.000	04:38	
Control Test	0.083	04:39	
Air Blank	0.000	04:39	
Subject Sample	# 1	04:39	Subject Test Refused
Breath Volume	0.000	04:39	
Air Blank	0.000	04:40	
Control Test	0.082	04:40	
Air Blank	0.000	04:41	
Diagnostics Check	OK	04:41	

Data Download Date/Time: 11/08/2018 11:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000782

INSTRUMENT REGISTERED TO: FHP TROOP A	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 770784
BREATH TEST OPERATOR: HELLETT,KARL,R	DRY GAS EXPIRY DATE: 04/27/2020
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)	VIOLATION CODE: DUI
ARRESTING OFFICER: HELLETT,KARL	ARREST TIME: 22:05
ARRESTING OFFICER AGENCY: FFWCC (Panama City)	GENDER: Male
LAST AGENCY INSPECTION: 08/22/2018	OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: W452103952410 DATE/TIME OF TEST: 09/12/2018 23:55:40

Diagnostics Check	OK	23:47
Air Blank	0.000	23:47
Control Test	0.080	23:47
Air Blank	0.000	23:48
Subject Sample # 1	0.025	23:51
Breath Volume	3.027	23:51
Air Blank	0.000	23:51
Air Blank	0.000	23:53
Subject Sample # 2	0.023	23:54
Breath Volume	3.703	23:54
Air Blank	0.000	23:54
Control Test	0.080	23:55
Air Blank	0.000	23:55
Diagnostics Check	OK	23:55

Data Download Date/Time: 09/14/2018 12:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN, MITCHELL, D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WHITE, DWAYNE
ARRESTING OFFICER AGENCY: Bonifay PD
LAST AGENCY INSPECTION: 09/14/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 461951
DRY GAS EXPIRY DATE: 06/15/2019
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: MILLS

DATE/TIME OF TEST: 09/22/2018 23:35:32

Diagnostics Check	OK	23:28	
Air Blank	0.000	23:29	
Control Test	0.080	23:29	
Air Blank	0.000	23:29	
Subject Sample # 1	0.171	23:30	
Breath Volume	3.640	23:30	
Air Blank	0.000	23:31	
Air Blank	0.000	23:33	
Subject Sample # 2	0.168	23:33	
Breath Volume	4.070	23:33	
Air Blank	0.000	23:34	
Control Test	0.074	23:34	Control Outside Tolerance
Air Blank	0.000	23:35	

Data Download Date/Time: 10/08/2018 21:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 770784
BREATH TEST OPERATOR: PYLE JR,HARRY,R	DRY GAS EXPIRY DATE: 04/27/2020
BREATH TEST OPERATOR AGENCY: FHP Troop A	VIOLATION CODE: DUI
ARRESTING OFFICER: PYLE JR,HARRY	ARREST TIME: 22:39
ARRESTING OFFICER AGENCY: FHP Troop A	GENDER: Male
LAST AGENCY INSPECTION: 08/22/2018	OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: JENKINS DATE/TIME OF TEST: 09/01/2018 00:14:59

Diagnostics Check	OK	00:07
Air Blank	0.000	00:08
Control Test	0.080	00:08
Air Blank	0.000	00:09
Subject Sample # 1	0.195	00:09
Breath Volume	2.757	00:09
Air Blank	0.000	00:10
Air Blank	0.000	00:12
Subject Sample # 2	0.197	00:13
Breath Volume	3.109	00:13
Air Blank	0.000	00:14
Control Test	0.079	00:14
Air Blank	0.000	00:14
Diagnostics Check	OK	00:14

Data Download Date/Time: 09/14/2018 12:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 770784
BREATH TEST OPERATOR: OPRY, WILLIAM, J	DRY GAS EXPIRY DATE: 04/27/2020
BREATH TEST OPERATOR AGENCY: FHP Troop A	VIOLATION CODE: DUI
ARRESTING OFFICER: WHITE, DWAYNE	ARREST TIME: 13:54
ARRESTING OFFICER AGENCY: Bonifay PD	GENDER: Male
LAST AGENCY INSPECTION: 08/22/2018	OBSERVATION PERIOD BEGAN: 13:54

SUBJECT'S LAST NAME: ZIMMERMAN	DATE/TIME OF TEST: 09/11/2018 14:33:45
--------------------------------	--

Diagnostics Check	OK	14:27
Air Blank	0.000	14:28
Control Test	0.079	14:28
Air Blank	0.000	14:29
Subject Sample # 1	0.019	14:29
Breath Volume	1.335	14:29
Air Blank	0.000	14:29
Air Blank	0.000	14:31
Subject Sample # 2	0.019	14:32
Breath Volume	1.269	14:32
Air Blank	0.000	14:32
Control Test	0.079	14:33
Air Blank	0.000	14:33
Diagnostics Check	OK	14:33

Data Download Date/Time: 09/14/2018 12:32

Subject Test Electronic Data

DATE/TIME OF TEST: 09/19/2018 21:15:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 770784
BREATH TEST OPERATOR: COX,DAVID,S	DRY GAS EXPIRY DATE: 04/27/2020
BREATH TEST OPERATOR AGENCY: FHP Troop A	VIOLATION CODE: DUI
ARRESTING OFFICER: COX,DAVID	ARREST TIME: 09:02
ARRESTING OFFICER AGENCY: FHP Troop A	GENDER: Female
LAST AGENCY INSPECTION: 09/14/2018	OBSERVATION PERIOD BEGAN: 11:27

SUBJECT'S LAST NAME: ROLISON DATE/TIME OF TEST: 09/24/2018 12:04:24

Diagnostics Check	OK	11:58
Air Blank	0.000	11:58
Control Test	0.081	11:59
Air Blank	0.000	11:59
Subject Sample # 1	0.000	12:00
Breath Volume	2.832	12:00
Air Blank	0.000	12:00
Air Blank	0.000	12:02
Subject Sample # 2	0.000	12:03
Breath Volume	2.695	12:03
Air Blank	0.000	12:03
Control Test	0.080	12:03
Air Blank	0.000	12:04
Diagnostics Check	OK	12:04

Data Download Date/Time: 10/20/2018 12:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TATRO,KERSTAN,J
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: TATRO,KERSTAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 08/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: SHIVELY

DATE/TIME OF TEST: 09/03/2018 03:46:04

Diagnostics Check	OK	03:38
Air Blank	0.000	03:39
Control Test	0.082	03:39
Air Blank	0.000	03:40
Subject Sample # 1	0.195	03:41
Breath Volume	1.457	03:41
Air Blank	0.000	03:41
Air Blank	0.000	03:43
Subject Sample # 2	0.185	03:44
Breath Volume	1.277	03:44
Air Blank	0.000	03:45
Control Test	0.082	03:45
Air Blank	0.000	03:45
Diagnostics Check	OK	03:46

Data Download Date/Time: 09/26/2018 15:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TATRO,KERSTAN,J
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: TATRO,KERSTAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 08/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: PICKLE

DATE/TIME OF TEST: 09/07/2018 05:09:31

Diagnostics Check	OK	05:02
Air Blank	0.000	05:02
Control Test	0.081	05:02
Air Blank	0.000	05:03
Subject Sample # 1	0.142	05:05
Breath Volume	2.734	05:05
Air Blank	0.000	05:05
Air Blank	0.000	05:07
Subject Sample # 2	0.144	05:08
Breath Volume	2.468	05:08
Air Blank	0.000	05:08
Control Test	0.080	05:08
Air Blank	0.000	05:09
Diagnostics Check	OK	05:09

Data Download Date/Time: 09/26/2018 15:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANKS,SEAN,C
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: BANKS,SEAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 08/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 09/09/2018 02:07:00

Diagnostics Check	OK	02:02	
Air Blank	0.000	02:02	
Control Test	0.081	02:03	
Air Blank	0.000	02:03	
Subject Sample	# 1	02:05	Subject Test Refused
Breath Volume	0.085	02:05	
Air Blank	0.000	02:06	
Control Test	0.081	02:06	
Air Blank	0.000	02:06	
Diagnostics Check	OK	02:07	

Data Download Date/Time: 09/26/2018 15:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANKS,SEAN,C
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: BANKS,SEAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 08/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 18:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:34

SUBJECT'S LAST NAME: INGRAHAM

DATE/TIME OF TEST: 09/18/2018 18:57:15

Diagnostics Check	OK	18:54	
Air Blank	0.000	18:54	
Control Test	0.082	18:54	
Air Blank	0.000	18:55	
Subject Sample	# 1	18:55	Subject Test Refused
Breath Volume	0.000	18:55	
Air Blank	0.000	18:56	
Control Test	0.081	18:56	
Air Blank	0.000	18:57	
Diagnostics Check	OK	18:57	

Data Download Date/Time: 09/26/2018 15:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNGREN,DENNIS,A
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: YOUNGREN,DENNIS
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 08/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: MEDINA

DATE/TIME OF TEST: 09/19/2018 01:58:26

Diagnostics Check	OK	01:55	
Air Blank	0.000	01:56	
Control Test	0.082	01:56	
Air Blank	0.000	01:56	
Subject Sample	# 1	01:57	RFI Detect
Breath Volume	0.000	01:57	
Air Blank	0.000	01:57	
Control Test	0.082	01:57	
Air Blank	0.000	01:58	
Diagnostics Check	OK	01:58	

Data Download Date/Time: 09/26/2018 15:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNGREN,DENNIS,A
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: YOUNGREN,DENNIS
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 08/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: MEDINA

DATE/TIME OF TEST: 09/19/2018 05:25:52

Diagnostics Check	OK	05:23	
Air Blank	0.000	05:23	
Control Test	0.082	05:23	
Air Blank	0.000	05:24	
Subject Sample	# 1	05:24	Subject Test Refused
Breath Volume	0.000	05:24	
Air Blank	0.000	05:24	
Control Test	0.082	05:25	
Air Blank	0.000	05:25	
Diagnostics Check	OK	05:25	

Data Download Date/Time: 09/26/2018 15:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNGREN,DENNIS,A
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: YOUNGREN,DENNIS
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 08/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 09/22/2018 02:04:34

Diagnostics Check	OK	01:58
Air Blank	0.000	01:58
Control Test	0.082	01:58
Air Blank	0.000	01:59
Subject Sample # 1	0.172	02:00
Breath Volume	2.121	02:00
Air Blank	0.000	02:00
Air Blank	0.000	02:02
Subject Sample # 2	0.169	02:03
Breath Volume	1.589	02:03
Air Blank	0.000	02:03
Control Test	0.082	02:03
Air Blank	0.000	02:04
Diagnostics Check	OK	02:04

Data Download Date/Time: 09/26/2018 15:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEIERS,BLAINE,T
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: WHITE,TRAVIS
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 08/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 17:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:19

SUBJECT'S LAST NAME: REALI

DATE/TIME OF TEST: 09/22/2018 17:46:01

Diagnostics Check	OK	17:39
Air Blank	0.000	17:40
Control Test	0.082	17:40
Air Blank	0.000	17:41
Subject Sample # 1	0.000	17:41
Breath Volume	1.421	17:41
Air Blank	0.000	17:42
Air Blank	0.000	17:44
Subject Sample # 2	0.000	17:44
Breath Volume	1.242	17:44
Air Blank	0.000	17:45
Control Test	0.082	17:45
Air Blank	0.000	17:45
Diagnostics Check	OK	17:46

Data Download Date/Time: 09/26/2018 15:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNGREN,DENNIS,A
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: YOUNGREN,DENNIS
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 08/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 09/22/2018 23:27:09

Diagnostics Check	OK	23:20	
Air Blank	0.000	23:21	
Control Test	0.082	23:21	
Air Blank	0.000	23:22	
Subject Sample # 1	0.000	23:22	
Breath Volume	3.789	23:22	
Air Blank	0.000	23:23	
Air Blank	0.000	23:25	
Subject Sample # 2	0.000	23:25	
Breath Volume	3.878	23:25	
Air Blank	0.000	23:26	
Control Test	0.082	23:26	
Air Blank		23:26	RFI Detect
Air Blank	0.000	23:27	

Data Download Date/Time: 09/26/2018 15:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANKS,SEAN,C
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: BANKS,SEAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 08/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: STROMAN

DATE/TIME OF TEST: 09/23/2018 02:52:15

Diagnostics Check	OK	02:49	
Air Blank	0.000	02:49	
Control Test	0.081	02:49	
Air Blank	0.000	02:50	
Subject Sample	# 1	02:50	Subject Test Refused
Breath Volume	0.000	02:50	
Air Blank	0.000	02:51	
Control Test	0.081	02:51	
Air Blank	0.000	02:52	
Diagnostics Check	OK	02:52	

Data Download Date/Time: 09/26/2018 15:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNGREN,DENNIS,A
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: YOUNGREN,DENNIS
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 08/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: COLSTON

DATE/TIME OF TEST: 09/24/2018 00:28:42

Diagnostics Check	OK	00:22
Air Blank	0.000	00:22
Control Test	0.082	00:23
Air Blank	0.000	00:23
Subject Sample # 1	0.134	00:24
Breath Volume	3.222	00:24
Air Blank	0.000	00:24
Air Blank	0.000	00:26
Subject Sample # 2	0.118	00:27
Breath Volume	1.125	00:27
Air Blank	0.000	00:27
Control Test	0.081	00:28
Air Blank	0.000	00:28
Diagnostics Check	OK	00:28

Data Download Date/Time: 09/26/2018 15:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000790

INSTRUMENT REGISTERED TO: VERO BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GASBARRINI,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: GASBARRINI,MICHAEL
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 08/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 13:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: DOE

DATE/TIME OF TEST: 09/07/2018 15:00:43

Diagnostics Check	OK	14:54
Air Blank	0.000	14:55
Control Test	0.078	14:55
Air Blank	0.000	14:56
Subject Sample # 1	0.000	14:56
Breath Volume	1.148	14:56
Air Blank	0.000	14:56
Air Blank	0.000	14:58
Subject Sample # 2	0.000	14:59
Breath Volume	2.851	14:59
Air Blank	0.000	14:59
Control Test	0.079	15:00
Air Blank	0.000	15:00
Diagnostics Check	OK	15:00

Data Download Date/Time: 09/30/2018 14:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000790

INSTRUMENT REGISTERED TO: VERO BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLUMB,GREGORY,T
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: PLUMB,GREGORY
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 08/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 09/07/2018 18:46:42

Diagnostics Check	OK	18:41
Air Blank	0.000	18:41
Control Test	0.080	18:41
Air Blank	0.000	18:42
Subject Sample # 1	0.000	18:42
Breath Volume	2.253	18:42
Air Blank	0.000	18:43
Air Blank	0.000	18:45
Subject Sample # 2	0.000	18:45
Breath Volume	1.812	18:45
Air Blank	0.000	18:45
Control Test	0.078	18:46
Air Blank	0.000	18:46
Diagnostics Check	OK	18:46

Data Download Date/Time: 09/30/2018 14:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000790

INSTRUMENT REGISTERED TO: VERO BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KERENSKY,BRIAN,M
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: ORTIZ,TAYLOR
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 08/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: MCGREGOR

DATE/TIME OF TEST: 09/14/2018 03:35:19

Diagnostics Check	OK	03:31	
Air Blank	0.000	03:32	
Control Test	0.079	03:32	
Air Blank	0.000	03:32	
Subject Sample	# 1	03:33	Subject Test Refused
Breath Volume	0.000	03:33	
Air Blank	0.000	03:34	
Control Test	0.080	03:34	
Air Blank	0.000	03:35	
Diagnostics Check	OK	03:35	

Data Download Date/Time: 09/30/2018 14:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: DRAWDY,LAINE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 09/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: MICELI

DATE/TIME OF TEST: 09/29/2018 01:45:48

Diagnostics Check	OK	01:39
Air Blank	0.000	01:39
Control Test	0.079	01:40
Air Blank	0.000	01:40
Subject Sample # 1	0.102	01:41
Breath Volume	1.718	01:41
Air Blank	0.000	01:41
Air Blank	0.000	01:43
Subject Sample # 2	0.098	01:44
Breath Volume	1.253	01:44
Air Blank	0.000	01:44
Control Test	0.079	01:45
Air Blank	0.000	01:45
Diagnostics Check	OK	01:45

Data Download Date/Time: 10/31/2018 08:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTON,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: ANTON,NICHOLAS
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 09/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: AYALA AREVALO

DATE/TIME OF TEST: 09/29/2018 04:33:09

Diagnostics Check	OK	04:26
Air Blank	0.000	04:27
Control Test	0.079	04:27
Air Blank	0.000	04:28
Subject Sample # 1	0.213	04:28
Breath Volume	1.441	04:28
Air Blank	0.000	04:29
Air Blank	0.000	04:31
Subject Sample # 2	0.216	04:31
Breath Volume	1.457	04:31
Air Blank	0.000	04:32
Control Test	0.081	04:32
Air Blank	0.000	04:33
Diagnostics Check	OK	04:33

Data Download Date/Time: 10/31/2018 08:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON,KYLE,L
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: SHERMETERO,MARTIN
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 08/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: BUI
ARREST TIME: 16:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: SPAGNUOLO

DATE/TIME OF TEST: 09/01/2018 18:01:20

Diagnostics Check	OK	17:57	
Air Blank	0.000	17:58	
Control Test	0.079	17:58	
Air Blank	0.000	17:58	
Subject Sample	# 1	17:59	Subject Test Refused
Breath Volume	0.000	17:59	
Air Blank	0.000	18:00	
Control Test	0.079	18:00	
Air Blank	0.000	18:01	
Diagnostics Check	OK	18:01	

Data Download Date/Time: 09/27/2018 08:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: HUNTER,RANDY
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 08/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: RICH

DATE/TIME OF TEST: 09/04/2018 01:08:32

Diagnostics Check	OK	01:02
Air Blank	0.000	01:02
Control Test	0.079	01:03
Air Blank	0.000	01:03
Subject Sample # 1	0.000	01:04
Breath Volume	1.929	01:04
Air Blank	0.000	01:04
Air Blank	0.000	01:06
Subject Sample # 2	0.000	01:07
Breath Volume	1.320	01:07
Air Blank	0.000	01:07
Control Test	0.079	01:07
Air Blank	0.000	01:08
Diagnostics Check	OK	01:08

Data Download Date/Time: 09/27/2018 08:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALTENBURG,MARCUS,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: RIDLE,EVAN
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 08/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:19

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 09/05/2018 22:08:06

Diagnostics Check	OK	21:58
Air Blank	0.000	21:58
Control Test	0.079	21:59
Air Blank	0.000	21:59
Subject Sample # 1	0.264	22:00
Breath Volume	2.062	22:00
Air Blank	0.000	22:01
Air Blank	0.000	22:02
Subject Sample # 2	0.243	22:03
Breath Volume	2.273	22:03
Air Blank	0.000	22:04
Air Blank	0.000	22:05
Subject Sample # 3	0.261	22:06
Breath Volume	1.761	22:06
Air Blank	0.000	22:07
Control Test	0.077	22:07
Air Blank	0.000	22:08
Diagnostics Check	OK	22:08

Data Download Date/Time: 09/27/2018 08:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: DRAWDY,LAINE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 08/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: HERNANDEZ HERNANDEZ

DATE/TIME OF TEST: 09/05/2018 23:20:48

Diagnostics Check	OK	23:14
Air Blank	0.000	23:15
Control Test	0.079	23:15
Air Blank	0.000	23:15
Subject Sample # 1	0.114	23:16
Breath Volume	1.750	23:16
Air Blank	0.000	23:17
Air Blank	0.000	23:18
Subject Sample # 2	0.117	23:19
Breath Volume	1.671	23:19
Air Blank	0.000	23:19
Control Test	0.078	23:20
Air Blank	0.000	23:20
Diagnostics Check	OK	23:20

Data Download Date/Time: 09/27/2018 08:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,RICHARD,D
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: HALL,RICHARD
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 08/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: J525552677090

DATE/TIME OF TEST: 09/07/2018 03:29:02

Diagnostics Check	OK	03:23	
Air Blank	0.000	03:24	
Control Test	0.079	03:24	
Air Blank	0.000	03:25	
Subject Sample	# 1	03:27	Subject Test Refused
Breath Volume	0.605	03:27	
Air Blank	0.000	03:28	
Control Test	0.079	03:28	
Air Blank	0.000	03:28	
Diagnostics Check	OK	03:29	

Data Download Date/Time: 09/27/2018 08:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: PATTERSON,JAMIE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 08/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: LEVERS

DATE/TIME OF TEST: 09/11/2018 00:35:52

Diagnostics Check	OK	00:28
Air Blank	0.000	00:29
Control Test	0.079	00:29
Air Blank	0.000	00:29
Subject Sample # 1	0.070	00:31
Breath Volume	1.406	00:31
Air Blank	0.000	00:31
Air Blank	0.000	00:33
Subject Sample # 2	0.068	00:34
Breath Volume	1.636	00:34
Air Blank	0.000	00:34
Control Test	0.078	00:35
Air Blank	0.000	00:35
Diagnostics Check	OK	00:35

Data Download Date/Time: 09/27/2018 08:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: DRAWDY,LAINE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 08/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: BOWERS

DATE/TIME OF TEST: 09/15/2018 01:10:09

Diagnostics Check	OK	01:03
Air Blank	0.000	01:04
Control Test	0.079	01:04
Air Blank	0.000	01:04
Subject Sample # 1	0.147	01:05
Breath Volume	1.269	01:05
Air Blank	0.000	01:05
Air Blank	0.000	01:07
Subject Sample # 2	0.145	01:08
Breath Volume	2.070	01:08
Air Blank	0.000	01:09
Control Test	0.080	01:09
Air Blank	0.000	01:10
Diagnostics Check	OK	01:10

Data Download Date/Time: 09/27/2018 08:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOHAMED,AHMED,S
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: ANDRADE,JESUS
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 08/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 12:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: MOHAMED

DATE/TIME OF TEST: 09/15/2018 01:44:50

Diagnostics Check	OK	01:32	
Air Blank	0.000	01:33	
Control Test	0.079	01:33	
Air Blank	0.000	01:33	
Subject Sample	# 1 0.000	01:36	Volume Not Met
Breath Volume	0.277	01:36	
Air Blank	0.000	01:37	
Air Blank	0.000	01:39	
Subject Sample	# 2 0.000	01:40	
Breath Volume	2.414	01:40	
Air Blank	0.000	01:40	
Air Blank	0.000	01:42	
Subject Sample	# 3 0.000	01:43	
Breath Volume	2.007	01:43	
Air Blank	0.000	01:43	
Control Test	0.079	01:44	
Air Blank	0.000	01:44	
Diagnostics Check	OK	01:44	

Data Download Date/Time: 09/27/2018 08:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: DRAWDY,LAINE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 08/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 09/15/2018 04:14:37

Diagnostics Check	OK	04:07
Air Blank	0.000	04:08
Control Test	0.079	04:08
Air Blank	0.000	04:09
Subject Sample # 1	0.131	04:10
Breath Volume	1.636	04:10
Air Blank	0.000	04:10
Air Blank	0.000	04:12
Subject Sample # 2	0.124	04:13
Breath Volume	1.902	04:13
Air Blank	0.000	04:13
Control Test	0.079	04:14
Air Blank	0.000	04:14
Diagnostics Check	OK	04:14

Data Download Date/Time: 09/27/2018 08:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLBERT,ERIC,R
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: HOLBERT,ERIC
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 08/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:18

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 09/17/2018 20:44:58

Diagnostics Check	OK	20:38
Air Blank	0.000	20:39
Control Test	0.079	20:39
Air Blank	0.000	20:40
Subject Sample # 1	0.193	20:40
Breath Volume	2.210	20:40
Air Blank	0.000	20:41
Air Blank	0.000	20:42
Subject Sample # 2	0.183	20:43
Breath Volume	1.765	20:43
Air Blank	0.000	20:44
Control Test	0.078	20:44
Air Blank	0.000	20:44
Diagnostics Check	OK	20:44

Data Download Date/Time: 09/27/2018 08:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: DRAWDY,LAINE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 08/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: BATES

DATE/TIME OF TEST: 09/20/2018 03:08:54

Diagnostics Check	OK	03:02
Air Blank	0.000	03:03
Control Test	0.080	03:03
Air Blank	0.000	03:03
Subject Sample # 1	0.150	03:04
Breath Volume	2.515	03:04
Air Blank	0.000	03:05
Air Blank	0.000	03:06
Subject Sample # 2	0.152	03:07
Breath Volume	2.382	03:07
Air Blank	0.000	03:07
Control Test	0.079	03:08
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

Data Download Date/Time: 09/27/2018 08:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALTENBURG,MARCUS,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: ETIENNE,EMMANUEL
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 08/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: HODGE

DATE/TIME OF TEST: 09/22/2018 01:48:39

Diagnostics Check	OK	01:42
Air Blank	0.000	01:42
Control Test	0.079	01:43
Air Blank	0.000	01:43
Subject Sample # 1	0.198	01:44
Breath Volume	1.867	01:44
Air Blank	0.000	01:44
Air Blank	0.000	01:46
Subject Sample # 2	0.189	01:47
Breath Volume	2.968	01:47
Air Blank	0.000	01:47
Control Test	0.077	01:48
Air Blank	0.000	01:48
Diagnostics Check	OK	01:48

Data Download Date/Time: 09/27/2018 08:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,TANYA,L
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: WRIGHT,TANYA
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 08/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: LOPEZ PERDOMO

DATE/TIME OF TEST: 09/22/2018 19:54:27

Diagnostics Check	OK	19:41	
Air Blank	0.000	19:42	
Control Test	0.079	19:42	
Air Blank	0.000	19:43	
Subject Sample	# 1 0.213	19:46	Volume Not Met
Breath Volume	1.000	19:46	
Air Blank	0.000	19:46	
Air Blank	0.000	19:48	
Subject Sample	# 2 0.223	19:49	No .020 agreement
Breath Volume	1.488	19:49	
Air Blank	0.000	19:50	
Air Blank	0.000	19:52	
Subject Sample	# 3 0.199	19:52	No .020 agreement
Breath Volume	1.238	19:52	
Air Blank	0.000	19:53	
Control Test	0.076	19:53	
Air Blank	0.000	19:54	
Diagnostics Check	OK	19:54	

Data Download Date/Time: 09/27/2018 08:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: DRAWDY,LAINE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 08/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: IZZO

DATE/TIME OF TEST: 09/25/2018 22:41:08

Diagnostics Check	OK	22:34
Air Blank	0.000	22:35
Control Test	0.080	22:35
Air Blank	0.000	22:35
Subject Sample # 1	0.053	22:36
Breath Volume	1.824	22:36
Air Blank	0.000	22:37
Air Blank	0.000	22:39
Subject Sample # 2	0.049	22:39
Breath Volume	1.136	22:39
Air Blank	0.000	22:40
Control Test	0.079	22:40
Air Blank	0.000	22:41
Diagnostics Check	OK	22:41

Data Download Date/Time: 09/27/2018 08:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000797

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALEXIS,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: HERNANDEZ,ALEXIS
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: TOSKE

DATE/TIME OF TEST: 09/05/2018 18:56:54

Diagnostics Check	OK	18:50
Air Blank	0.000	18:51
Control Test	0.081	18:51
Air Blank	0.000	18:51
Subject Sample # 1	0.204	18:52
Breath Volume	1.636	18:52
Air Blank	0.000	18:52
Air Blank	0.000	18:54
Subject Sample # 2	0.188	18:55
Breath Volume	1.554	18:55
Air Blank	0.000	18:56
Control Test	0.080	18:56
Air Blank	0.000	18:56
Diagnostics Check	OK	18:56

Data Download Date/Time: 09/29/2018 22:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000809

INSTRUMENT REGISTERED TO: DIXIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKELLY,K.,E
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SKELLY,K.
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: KATSAROS

DATE/TIME OF TEST: 09/02/2018 01:33:31

Diagnostics Check	OK	01:30	
Air Blank	0.000	01:31	
Control Test	0.080	01:31	
Air Blank	0.000	01:31	
Subject Sample	# 1	01:32	Subject Test Refused
Breath Volume	0.000	01:32	
Air Blank	0.000	01:32	
Control Test	0.080	01:32	
Air Blank	0.000	01:33	
Diagnostics Check	OK	01:33	

Data Download Date/Time: 09/28/2018 13:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000809

INSTRUMENT REGISTERED TO: DIXIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FUTCH,JAMES,M
BREATH TEST OPERATOR AGENCY: Dixie County SO
ARRESTING OFFICER: SINGLETON,NATHAN
ARRESTING OFFICER AGENCY: Dixie County SO
LAST AGENCY INSPECTION: 08/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:18

SUBJECT'S LAST NAME: FOLSOM

DATE/TIME OF TEST: 09/09/2018 18:13:19

Diagnostics Check	OK	18:07
Air Blank	0.000	18:07
Control Test	0.080	18:08
Air Blank	0.000	18:08
Subject Sample # 1	0.294	18:09
Breath Volume	1.335	18:09
Air Blank	0.000	18:09
Air Blank	0.000	18:11
Subject Sample # 2	0.301	18:11
Breath Volume	1.589	18:11
Air Blank	0.000	18:12
Control Test	0.079	18:12
Air Blank	0.000	18:13
Diagnostics Check	OK	18:13

Data Download Date/Time: 09/28/2018 13:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000809

INSTRUMENT REGISTERED TO: DIXIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHRADER, WILLIAM, F
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SINGLETON, NATHAN
ARRESTING OFFICER AGENCY: Dixie County SO
LAST AGENCY INSPECTION: 08/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:33

SUBJECT'S LAST NAME: JORDAN

DATE/TIME OF TEST: 09/21/2018 20:58:15

Diagnostics Check	OK	20:57	
Air Blank		20:57	RFI Detect
Air Blank	0.000	20:58	

Data Download Date/Time: 09/28/2018 13:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000809

INSTRUMENT REGISTERED TO: DIXIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHRADER, WILLIAM, F
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SINGLETON, NATHAN
ARRESTING OFFICER AGENCY: Dixie County SO
LAST AGENCY INSPECTION: 08/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:33

SUBJECT'S LAST NAME: JORDAN

DATE/TIME OF TEST: 09/21/2018 21:06:25

Diagnostics Check	OK	21:00
Air Blank	0.000	21:00
Control Test	0.080	21:01
Air Blank	0.000	21:01
Subject Sample # 1	0.148	21:02
Breath Volume	2.453	21:02
Air Blank	0.000	21:02
Air Blank	0.000	21:04
Subject Sample # 2	0.145	21:04
Breath Volume	3.222	21:04
Air Blank	0.000	21:05
Control Test	0.079	21:05
Air Blank	0.000	21:06
Diagnostics Check	OK	21:06

Data Download Date/Time: 09/28/2018 13:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000809

INSTRUMENT REGISTERED TO: DIXIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FULTON,DARIN,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SINGLETON,NATHAN
ARRESTING OFFICER AGENCY: Dixie County SO
LAST AGENCY INSPECTION: 08/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: HURST

DATE/TIME OF TEST: 09/24/2018 01:19:36

Diagnostics Check	OK	01:12
Air Blank	0.000	01:13
Control Test	0.079	01:13
Air Blank	0.000	01:14
Subject Sample # 1	0.105	01:15
Breath Volume	1.496	01:15
Air Blank	0.000	01:15
Air Blank	0.000	01:17
Subject Sample # 2	0.113	01:18
Breath Volume	1.679	01:18
Air Blank	0.000	01:18
Control Test	0.080	01:19
Air Blank	0.000	01:19
Diagnostics Check	OK	01:19

Data Download Date/Time: 09/28/2018 13:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000811

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CEPHUS,JAMIN,I
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: CEPHUS,JAMIN
ARRESTING OFFICER AGENCY: Perry PD
LAST AGENCY INSPECTION: 08/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 09/08/2018 22:36:21

Diagnostics Check	OK	22:29
Air Blank	0.000	22:29
Control Test	0.077	22:30
Air Blank	0.000	22:30
Subject Sample # 1	0.000	22:31
Breath Volume	1.656	22:31
Air Blank	0.000	22:32
Air Blank	0.000	22:34
Subject Sample # 2	0.000	22:35
Breath Volume	1.296	22:35
Air Blank	0.000	22:35
Control Test	0.076	22:35
Air Blank	0.000	22:36
Diagnostics Check	OK	22:36

Data Download Date/Time: 12/12/2018 14:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000811

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHELPS, WILLIAM, D
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: DICE, RICKY
ARRESTING OFFICER AGENCY: Perry PD
LAST AGENCY INSPECTION: 08/29/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: HARDAN

DATE/TIME OF TEST: 09/23/2018 00:18:57

Diagnostics Check	OK	00:12
Air Blank	0.000	00:13
Control Test	0.076	00:13
Air Blank	0.000	00:13
Subject Sample # 1	0.191	00:14
Breath Volume	1.726	00:14
Air Blank	0.000	00:14
Air Blank	0.000	00:16
Subject Sample # 2	0.189	00:17
Breath Volume	2.054	00:17
Air Blank	0.000	00:18
Control Test	0.075	00:18
Air Blank	0.000	00:18
Diagnostics Check	OK	00:18

Data Download Date/Time: 12/12/2018 14:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000811

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BASS,JASON,P
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: SWINDLE,CHARLES
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 08/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 14:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:03

SUBJECT'S LAST NAME: DICE

DATE/TIME OF TEST: 09/24/2018 15:30:19

Diagnostics Check	OK	15:23
Air Blank	0.000	15:24
Control Test	0.076	15:24
Air Blank	0.000	15:24
Subject Sample # 1	0.000	15:25
Breath Volume	3.187	15:25
Air Blank	0.000	15:26
Air Blank	0.000	15:28
Subject Sample # 2	0.000	15:28
Breath Volume	2.835	15:28
Air Blank	0.000	15:29
Control Test	0.075	15:29
Air Blank	0.000	15:30
Diagnostics Check	OK	15:30

Data Download Date/Time: 12/12/2018 14:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLIS,TERENCE,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: GREATREX,T
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 09/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:02

SUBJECT'S LAST NAME: BRUZZESE

DATE/TIME OF TEST: 09/27/2018 21:31:51

Diagnostics Check	OK	21:22
Air Blank	0.000	21:23
Control Test	0.078	21:23
Air Blank	0.000	21:24
Subject Sample # 1	0.181	21:25
Breath Volume	2.402	21:25
Air Blank	0.000	21:25
Air Blank	0.000	21:27
Subject Sample # 2	0.178	21:30
Breath Volume	2.417	21:30
Air Blank	0.000	21:30
Control Test	0.079	21:31
Air Blank	0.000	21:31
Diagnostics Check	OK	21:31

Data Download Date/Time: 10/28/2018 10:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONG,ERIC,W
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: STEMEN,COREY
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 08/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:47

SUBJECT'S LAST NAME: BERRIOS

DATE/TIME OF TEST: 09/02/2018 05:16:56

Diagnostics Check	OK	05:09
Air Blank	0.000	05:10
Control Test	0.080	05:10
Air Blank	0.000	05:11
Subject Sample # 1	0.172	05:11
Breath Volume	1.347	05:11
Air Blank	0.000	05:12
Air Blank	0.000	05:14
Subject Sample # 2	0.159	05:15
Breath Volume	1.195	05:15
Air Blank	0.000	05:15
Control Test	0.079	05:16
Air Blank	0.000	05:16
Diagnostics Check	OK	05:16

Data Download Date/Time: 09/27/2018 09:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONG,ERIC,W
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: NORDLE,LACEY
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 08/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: ENDICOTT

DATE/TIME OF TEST: 09/03/2018 01:36:06

Diagnostics Check	OK	01:29
Air Blank	0.000	01:29
Control Test	0.080	01:29
Air Blank	0.000	01:30
Subject Sample # 1	0.237	01:31
Breath Volume	1.199	01:31
Air Blank	0.000	01:32
Air Blank	0.000	01:33
Subject Sample # 2	0.217	01:34
Breath Volume	1.214	01:34
Air Blank	0.000	01:35
Control Test	0.078	01:35
Air Blank	0.000	01:36
Diagnostics Check	OK	01:36

Data Download Date/Time: 09/27/2018 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLIS,TERENCE,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: BOUTROUS,P
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 08/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: OLIVO JR

DATE/TIME OF TEST: 09/05/2018 01:46:25

Diagnostics Check	OK	01:43	
Air Blank	0.000	01:43	
Control Test	0.080	01:44	
Air Blank	0.000	01:44	
Subject Sample	# 1	01:45	Subject Test Refused
Breath Volume	0.000	01:45	
Air Blank	0.000	01:45	
Control Test	0.079	01:45	
Air Blank	0.000	01:46	
Diagnostics Check	OK	01:46	

Data Download Date/Time: 09/27/2018 09:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLIS,TERENCE,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: GREATREX,T
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 08/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: FERKO JR

DATE/TIME OF TEST: 09/07/2018 22:25:23

Diagnostics Check	OK	22:22	
Air Blank	0.000	22:22	
Control Test	0.079	22:23	
Air Blank	0.000	22:23	
Subject Sample	# 1	22:23	Subject Test Refused
Breath Volume	0.000	22:23	
Air Blank	0.000	22:24	
Control Test	0.079	22:24	
Air Blank	0.000	22:25	
Diagnostics Check	OK	22:25	

Data Download Date/Time: 09/27/2018 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLIS,TERENCE,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: STEWART,D
ARRESTING OFFICER AGENCY: Coleman PD
LAST AGENCY INSPECTION: 08/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: ADALBERTO

DATE/TIME OF TEST: 09/08/2018 23:40:44

Diagnostics Check	OK	23:37	
Air Blank	0.000	23:38	
Control Test	0.079	23:38	
Air Blank	0.000	23:39	
Subject Sample	# 1	23:39	Subject Test Refused
Breath Volume	0.000	23:39	
Air Blank	0.000	23:39	
Control Test	0.080	23:40	
Air Blank	0.000	23:40	
Diagnostics Check	OK	23:40	

Data Download Date/Time: 09/27/2018 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,JARROD,W
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: NELSON,GARRETT
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 08/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 12:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:21

SUBJECT'S LAST NAME: RIMMER

DATE/TIME OF TEST: 09/09/2018 14:47:47

Diagnostics Check	OK	14:41
Air Blank	0.000	14:41
Control Test	0.079	14:41
Air Blank	0.000	14:42
Subject Sample # 1	0.150	14:42
Breath Volume	1.492	14:42
Air Blank	0.000	14:43
Air Blank	0.000	14:45
Subject Sample # 2	0.167	14:46
Breath Volume	1.437	14:46
Air Blank	0.000	14:46
Control Test	0.080	14:47
Air Blank	0.000	14:47
Diagnostics Check	OK	14:47

Data Download Date/Time: 09/27/2018 09:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,JARROD,W
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: BOUTROS,PHILLIP
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 08/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:22

SUBJECT'S LAST NAME: BALDWIN

DATE/TIME OF TEST: 09/13/2018 16:49:42

Diagnostics Check	OK	16:43
Air Blank	0.000	16:43
Control Test	0.080	16:44
Air Blank	0.000	16:44
Subject Sample # 1	0.000	16:45
Breath Volume	1.339	16:45
Air Blank	0.000	16:45
Air Blank	0.000	16:47
Subject Sample # 2	0.000	16:48
Breath Volume	1.246	16:48
Air Blank	0.000	16:48
Control Test	0.080	16:49
Air Blank	0.000	16:49
Diagnostics Check	OK	16:49

Data Download Date/Time: 09/27/2018 09:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONG,ERIC,W
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: ZINN,BRYAN
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 08/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 09/15/2018 00:40:37

Diagnostics Check	OK	00:33
Air Blank	0.000	00:33
Control Test	0.079	00:34
Air Blank	0.000	00:34
Subject Sample # 1	0.167	00:35
Breath Volume	1.648	00:35
Air Blank	0.000	00:36
Air Blank	0.000	00:38
Subject Sample # 2	0.164	00:38
Breath Volume	1.472	00:38
Air Blank	0.000	00:39
Control Test	0.078	00:39
Air Blank	0.000	00:40
Diagnostics Check	OK	00:40

Data Download Date/Time: 09/27/2018 09:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONG,ERIC,W
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: ELIAS,DANIEL
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 08/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:27

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 09/15/2018 20:10:17

Diagnostics Check	OK	20:02
Air Blank	0.000	20:03
Control Test	0.080	20:03
Air Blank	0.000	20:04
Subject Sample # 1	0.148	20:05
Breath Volume	2.457	20:05
Air Blank	0.000	20:06
Air Blank	0.000	20:07
Subject Sample # 2	0.144	20:08
Breath Volume	1.734	20:08
Air Blank	0.000	20:09
Control Test	0.080	20:09
Air Blank	0.000	20:10
Diagnostics Check	OK	20:10

Data Download Date/Time: 09/27/2018 09:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLIS,TERENCE,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: RAY,A
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 08/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: CRAWFORD

DATE/TIME OF TEST: 09/19/2018 02:04:17

Diagnostics Check	OK	02:01	
Air Blank	0.000	02:01	
Control Test	0.080	02:02	
Air Blank	0.000	02:02	
Subject Sample	# 1	02:02	Subject Test Refused
Breath Volume	0.000	02:02	
Air Blank	0.000	02:03	
Control Test	0.078	02:03	
Air Blank	0.000	02:04	
Diagnostics Check	OK	02:04	

Data Download Date/Time: 09/27/2018 09:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLIS,TERENCE,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: SCHMIDT,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 09/23/2018 00:04:57

Diagnostics Check	OK	23:58
Air Blank	0.000	23:58
Control Test	0.079	23:58
Air Blank	0.000	23:59
Subject Sample # 1	0.157	00:00
Breath Volume	2.082	00:00
Air Blank	0.000	00:00
Air Blank	0.000	00:02
Subject Sample # 2	0.154	00:03
Breath Volume	2.097	00:03
Air Blank	0.000	00:03
Control Test	0.079	00:04
Air Blank	0.000	00:04
Diagnostics Check	OK	00:04

Data Download Date/Time: 09/27/2018 09:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLIS,TERENCE,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: DOBSIEWICZ,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/28/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:36

SUBJECT'S LAST NAME: RAGSDALE

DATE/TIME OF TEST: 09/24/2018 06:04:05

Diagnostics Check	OK	05:57
Air Blank	0.000	05:57
Control Test	0.079	05:57
Air Blank	0.000	05:58
Subject Sample # 1	0.242	05:59
Breath Volume	4.019	05:59
Air Blank	0.000	05:59
Air Blank	0.000	06:01
Subject Sample # 2	0.233	06:02
Breath Volume	2.785	06:02
Air Blank	0.000	06:03
Control Test	0.078	06:03
Air Blank	0.000	06:04
Diagnostics Check	OK	06:04

Data Download Date/Time: 09/27/2018 09:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONG,ERIC,W
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: RAFFERTY,CHRIS
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 08/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 09/26/2018 00:25:47

Diagnostics Check	OK	00:22	
Air Blank	0.000	00:22	
Control Test	0.079	00:23	
Air Blank	0.000	00:23	
Subject Sample	# 1	00:24	Subject Test Refused
Breath Volume	0.000	00:24	
Air Blank	0.000	00:24	
Control Test	0.079	00:25	
Air Blank	0.000	00:25	
Diagnostics Check	OK	00:25	

Data Download Date/Time: 09/27/2018 09:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLIS,TERENCE,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: BOUTRAS,P
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 08/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: SPANN

DATE/TIME OF TEST: 09/26/2018 19:19:31

Diagnostics Check	OK	19:10	
Air Blank	0.000	19:10	
Control Test	0.079	19:11	
Air Blank	0.000	19:11	
Subject Sample	# 1 0.032	19:14	Volume Not Met
Breath Volume	0.261	19:14	
Air Blank	0.000	19:15	
Air Blank	0.000	19:17	
Subject Sample	# 2	19:18	Subject Test Refused
Breath Volume	0.000	19:18	
Air Blank	0.000	19:18	
Control Test	0.079	19:18	
Air Blank	0.000	19:19	
Diagnostics Check	OK	19:19	

Data Download Date/Time: 09/27/2018 09:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,CHASE,A
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: KJEKSTAD,ERIC
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: ROBBINS

DATE/TIME OF TEST: 09/01/2018 23:17:45

Diagnostics Check	OK	23:10
Air Blank	0.000	23:10
Control Test	0.078	23:10
Air Blank	0.000	23:11
Subject Sample # 1	0.232	23:11
Breath Volume	1.121	23:11
Air Blank	0.000	23:12
Air Blank	0.000	23:14
Subject Sample # 2	0.223	23:16
Breath Volume	1.253	23:16
Air Blank	0.000	23:16
Control Test	0.078	23:17
Air Blank	0.000	23:17
Diagnostics Check	OK	23:17

Data Download Date/Time: 09/24/2018 14:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLOYD,DANIEL,Z
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: GERENA,DEP
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 08/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: BARNETT

DATE/TIME OF TEST: 09/04/2018 02:28:23

Diagnostics Check	OK	02:25	
Air Blank	0.000	02:25	
Control Test	0.078	02:26	
Air Blank	0.000	02:26	
Subject Sample	# 1	02:27	Subject Test Refused
Breath Volume	0.000	02:27	
Air Blank	0.000	02:27	
Control Test	0.079	02:27	
Air Blank	0.000	02:28	
Diagnostics Check	OK	02:28	

Data Download Date/Time: 09/24/2018 14:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLOYD,DANIEL,Z
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: WARD,TROOPER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: VOGEL

DATE/TIME OF TEST: 09/08/2018 04:38:47

Diagnostics Check	OK	04:31
Air Blank	0.000	04:32
Control Test	0.078	04:32
Air Blank	0.000	04:33
Subject Sample # 1	0.266	04:34
Breath Volume	2.218	04:34
Air Blank	0.000	04:34
Air Blank	0.000	04:36
Subject Sample # 2	0.254	04:37
Breath Volume	1.687	04:37
Air Blank	0.000	04:37
Control Test	0.077	04:38
Air Blank	0.000	04:38
Diagnostics Check	OK	04:38

Data Download Date/Time: 09/24/2018 14:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLOYD,DANIEL,Z
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: RAINES,OFFICER
ARRESTING OFFICER AGENCY: Dunnellon PD
LAST AGENCY INSPECTION: 08/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 09/09/2018 04:24:11

Diagnostics Check	OK	04:12	
Air Blank	0.000	04:13	
Control Test	0.079	04:13	
Air Blank	0.000	04:13	
Subject Sample	# 1	04:17	No Sample Provided
Breath Volume	0.000	04:17	
Air Blank	0.000	04:17	
Air Blank	0.000	04:19	
Subject Sample	# 2	0.121	04:22
Breath Volume	0.656	04:22	Volume Not Met
Air Blank	0.000	04:23	
Control Test	0.079	04:23	
Air Blank	0.000	04:24	
Diagnostics Check	OK	04:24	

Data Download Date/Time: 09/24/2018 14:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLOYD,DANIEL,Z
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: RAINES,OFFICER
ARRESTING OFFICER AGENCY: Dunnellon PD
LAST AGENCY INSPECTION: 08/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 09/09/2018 04:37:30

Diagnostics Check	OK	04:26
Air Blank	0.000	04:27
Control Test	0.078	04:27
Air Blank	0.000	04:27
Subject Sample # 1	0.143	04:30
Breath Volume	1.945	04:30
Air Blank	0.000	04:30
Air Blank	0.000	04:32
Subject Sample # 2	0.138	04:35
Breath Volume	2.132	04:35
Air Blank	0.000	04:36
Control Test	0.079	04:36
Air Blank	0.000	04:37
Diagnostics Check	OK	04:37

Data Download Date/Time: 09/24/2018 14:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WING,JOSHUA,M
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: COUNTY WORKER,COUNTY
WORKER
ARRESTING OFFICER AGENCY: CCA Citrus
LAST AGENCY INSPECTION: 08/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: ZERO TOL
ARREST TIME: 00:00
GENDER: Male

OBSERVATION PERIOD BEGAN: 11:45

SUBJECT'S LAST NAME: HIRSCH

DATE/TIME OF TEST: 09/13/2018 12:12:28

Diagnostics Check	OK	12:05
Air Blank	0.000	12:06
Control Test	0.079	12:06
Air Blank	0.000	12:07
Subject Sample # 1	0.000	12:07
Breath Volume	2.144	12:07
Air Blank	0.000	12:08
Air Blank	0.000	12:10
Subject Sample # 2	0.000	12:11
Breath Volume	2.253	12:11
Air Blank	0.000	12:11
Control Test	0.079	12:11
Air Blank	0.000	12:12
Diagnostics Check	OK	12:12

Data Download Date/Time: 09/24/2018 14:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WING,JOSHUA,M
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: COUNTY WORKER,COUNTY
WORKER
ARRESTING OFFICER AGENCY: CCA Citrus
LAST AGENCY INSPECTION: 08/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: ZERO TOL
ARREST TIME: 00:00
GENDER: Male

OBSERVATION PERIOD BEGAN: 12:06

SUBJECT'S LAST NAME: HIRSCH

DATE/TIME OF TEST: 09/13/2018 12:32:39

Diagnostics Check	OK	12:26
Air Blank	0.000	12:26
Control Test	0.078	12:27
Air Blank	0.000	12:27
Subject Sample # 1	0.000	12:28
Breath Volume	2.250	12:28
Air Blank	0.000	12:28
Air Blank	0.000	12:30
Subject Sample # 2	0.000	12:31
Breath Volume	2.207	12:31
Air Blank	0.000	12:31
Control Test	0.078	12:32
Air Blank	0.000	12:32
Diagnostics Check	OK	12:32

Data Download Date/Time: 09/24/2018 14:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WING,JOSHUA,M
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: BARD,TROOPER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 14:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:37

SUBJECT'S LAST NAME: LIESNER

DATE/TIME OF TEST: 09/13/2018 16:04:08

Diagnostics Check	OK	15:57
Air Blank	0.000	15:58
Control Test	0.078	15:58
Air Blank	0.000	15:59
Subject Sample # 1	0.205	15:59
Breath Volume	2.546	15:59
Air Blank	0.000	16:00
Air Blank	0.000	16:02
Subject Sample # 2	0.206	16:02
Breath Volume	2.484	16:02
Air Blank	0.000	16:03
Control Test	0.078	16:03
Air Blank	0.000	16:04
Diagnostics Check	OK	16:04

Data Download Date/Time: 09/24/2018 14:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,CHASE,A
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: BROWNING,JON
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 08/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: FOWLER

DATE/TIME OF TEST: 09/15/2018 03:23:44

Diagnostics Check	OK	03:17
Air Blank	0.000	03:18
Control Test	0.079	03:18
Air Blank	0.000	03:18
Subject Sample # 1	0.246	03:19
Breath Volume	1.753	03:19
Air Blank	0.000	03:19
Air Blank	0.000	03:21
Subject Sample # 2	0.248	03:22
Breath Volume	2.515	03:22
Air Blank	0.000	03:22
Control Test	0.078	03:23
Air Blank	0.000	03:23
Diagnostics Check	OK	03:23

Data Download Date/Time: 09/24/2018 14:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWIGER,CARL,M
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: HOLTZHOUSE,N/A
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 08/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 11:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:15

SUBJECT'S LAST NAME: JOHNSTON

DATE/TIME OF TEST: 09/16/2018 12:43:18

Diagnostics Check	OK	12:40	
Air Blank	0.000	12:40	
Control Test	0.079	12:41	
Air Blank	0.000	12:41	
Subject Sample	# 1	12:41	Subject Test Refused
Breath Volume	0.000	12:41	
Air Blank	0.000	12:42	
Control Test	0.079	12:42	
Air Blank	0.000	12:43	
Diagnostics Check	OK	12:43	

Data Download Date/Time: 09/24/2018 14:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLOYD,DANIEL,Z
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: TAYLOR,T
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 08/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: WOLFE

DATE/TIME OF TEST: 09/18/2018 22:40:22

Diagnostics Check	OK	22:32
Air Blank	0.000	22:33
Control Test	0.078	22:33
Air Blank	0.000	22:34
Subject Sample # 1	0.175	22:35
Breath Volume	1.140	22:35
Air Blank	0.000	22:36
Air Blank	0.000	22:37
Subject Sample # 2	0.178	22:38
Breath Volume	2.039	22:38
Air Blank	0.000	22:39
Control Test	0.077	22:39
Air Blank	0.000	22:40
Diagnostics Check	OK	22:40

Data Download Date/Time: 09/24/2018 14:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLOYD,DANIEL,Z
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: PASTERNAK,TROOPER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:47

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 09/20/2018 20:10:48

Diagnostics Check	OK	20:07	
Air Blank	0.000	20:08	
Control Test	0.078	20:08	
Air Blank	0.000	20:09	
Subject Sample	# 1	20:09	Subject Test Refused
Breath Volume	0.000	20:09	
Air Blank	0.000	20:09	
Control Test	0.078	20:10	
Air Blank	0.000	20:10	
Diagnostics Check	OK	20:10	

Data Download Date/Time: 09/24/2018 14:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,CHASE,A
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: LAKATIS,K
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 09/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 09/30/2018 01:44:44

Diagnostics Check	OK	01:37
Air Blank	0.000	01:37
Control Test	0.078	01:38
Air Blank	0.000	01:38
Subject Sample # 1	0.220	01:39
Breath Volume	3.269	01:39
Air Blank	0.000	01:40
Air Blank	0.000	01:42
Subject Sample # 2	0.222	01:43
Breath Volume	2.714	01:43
Air Blank	0.000	01:43
Control Test	0.078	01:44
Air Blank	0.000	01:44
Diagnostics Check	OK	01:44

Data Download Date/Time: 11/05/2018 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,CHASE,A
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: WATTS,R
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 09/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: HOLTON

DATE/TIME OF TEST: 09/30/2018 04:16:45

Diagnostics Check	OK	04:04	
Air Blank	0.000	04:04	
Control Test	0.078	04:04	
Air Blank	0.000	04:05	
Subject Sample	# 1 0.119	04:08	Volume Not Met
Breath Volume	0.062	04:08	
Air Blank	0.000	04:09	
Air Blank	0.000	04:10	
Subject Sample	# 2 0.135	04:12	
Breath Volume	1.675	04:12	
Air Blank	0.000	04:12	
Air Blank	0.000	04:14	
Subject Sample	# 3 0.127	04:15	
Breath Volume	1.886	04:15	
Air Blank	0.000	04:15	
Control Test	0.077	04:16	
Air Blank	0.000	04:16	
Diagnostics Check	OK	04:16	

Data Download Date/Time: 11/05/2018 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHESSHER,SANDRA,W
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: HICKMAN,JAMAAL
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 08/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: LANSRUD

DATE/TIME OF TEST: 09/01/2018 02:52:25

Diagnostics Check	OK	02:45
Air Blank	0.000	02:45
Control Test	0.076	02:46
Air Blank	0.000	02:46
Subject Sample # 1	0.139	02:47
Breath Volume	2.089	02:47
Air Blank	0.000	02:48
Air Blank	0.000	02:50
Subject Sample # 2	0.143	02:50
Breath Volume	2.339	02:50
Air Blank	0.000	02:51
Control Test	0.076	02:51
Air Blank	0.000	02:52
Diagnostics Check	OK	02:52

Data Download Date/Time: 09/04/2018 13:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

INSTRUMENT REGISTERED TO: MOUNT DORA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICE,A,
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: BLACKWELL,D
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 08/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 583457
DRY GAS EXPIRY DATE: 10/20/2019
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 09/20/2018 22:32:22

Diagnostics Check	OK	22:29	
Air Blank	0.000	22:29	
Control Test	0.076	22:30	
Air Blank	0.000	22:30	
Subject Sample	# 1	22:31	Subject Test Refused
Breath Volume	0.000	22:31	
Air Blank	0.000	22:31	
Control Test	0.077	22:31	
Air Blank	0.000	22:32	
Diagnostics Check	OK	22:32	

Data Download Date/Time: 09/30/2018 06:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LETSON,REBECCA,G
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: LETSON,REBECCA
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 08/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: KELLER

DATE/TIME OF TEST: 09/01/2018 03:41:25

Diagnostics Check	OK	03:40	
Air Blank		03:40	RFI Detect
Air Blank	0.000	03:41	

Data Download Date/Time: 09/29/2018 12:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LETSON,REBECCA,G
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: LETSON,REBECCA
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 08/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: KELLER

DATE/TIME OF TEST: 09/01/2018 03:49:58

Diagnostics Check	OK	03:43
Air Blank	0.000	03:43
Control Test	0.080	03:43
Air Blank	0.000	03:44
Subject Sample # 1	0.210	03:45
Breath Volume	1.628	03:45
Air Blank	0.000	03:45
Air Blank	0.000	03:47
Subject Sample # 2	0.200	03:48
Breath Volume	1.687	03:48
Air Blank	0.000	03:49
Control Test	0.079	03:49
Air Blank	0.000	03:49
Diagnostics Check	OK	03:49

Data Download Date/Time: 09/29/2018 12:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARTMAN,RILEY,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: ARTMAN,RILEY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: HUTTO

DATE/TIME OF TEST: 09/03/2018 01:58:45

Diagnostics Check	OK	01:42	
Air Blank	0.000	01:42	
Control Test	0.081	01:43	
Air Blank	0.000	01:43	
Subject Sample	# 1 0.076	01:45	
Breath Volume	2.453	01:45	
Air Blank	0.000	01:46	
Air Blank	0.000	01:48	
Subject Sample	# 2 0.054	01:51	Volume Not Met
Breath Volume	0.148	01:51	
Air Blank	0.000	01:52	
Air Blank	0.000	01:53	
Subject Sample	# 3 0.067	01:57	Volume Not Met
Breath Volume	0.191	01:57	
Air Blank	0.000	01:57	
Control Test	0.078	01:58	
Air Blank	0.000	01:58	
Diagnostics Check	OK	01:58	

Data Download Date/Time: 09/29/2018 12:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRINGTON,ALEX,M
BREATH TEST OPERATOR AGENCY: Belleview PD
ARRESTING OFFICER: HARRINGTON,ALEX
ARRESTING OFFICER AGENCY: Belleview PD
LAST AGENCY INSPECTION: 08/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: HUNTER

DATE/TIME OF TEST: 09/04/2018 00:59:48

Diagnostics Check	OK	00:53
Air Blank	0.000	00:53
Control Test	0.081	00:54
Air Blank	0.000	00:54
Subject Sample # 1	0.039	00:55
Breath Volume	1.832	00:55
Air Blank	0.000	00:55
Air Blank	0.000	00:57
Subject Sample # 2	0.037	00:58
Breath Volume	1.203	00:58
Air Blank	0.000	00:58
Control Test	0.081	00:59
Air Blank	0.000	00:59
Diagnostics Check	OK	00:59

Data Download Date/Time: 09/29/2018 12:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOODY,PATRICK,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: DOODY,PATRICK
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 13:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:15

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 09/06/2018 14:57:52

Diagnostics Check	OK	14:44	
Air Blank	0.000	14:45	
Control Test	0.081	14:45	
Air Blank	0.000	14:45	
Subject Sample	# 1 0.000	14:49	Volume Not Met
Breath Volume	0.585	14:49	
Air Blank	0.000	14:49	
Air Blank	0.000	14:51	
Subject Sample	# 2 0.000	14:53	
Breath Volume	2.078	14:53	
Air Blank	0.000	14:53	
Air Blank	0.000	14:55	
Subject Sample	# 3 0.000	14:56	
Breath Volume	1.445	14:56	
Air Blank	0.000	14:56	
Control Test	0.078	14:57	
Air Blank	0.000	14:57	
Diagnostics Check	OK	14:57	

Data Download Date/Time: 09/29/2018 12:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: UMANA,RAUL,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: UMANA,RAUL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 10:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: BOLAND

DATE/TIME OF TEST: 09/06/2018 22:58:03

Diagnostics Check	OK	22:52	
Air Blank	0.000	22:52	
Control Test	0.081	22:52	
Air Blank	0.000	22:53	
Subject Sample	# 1 0.158	22:54	
Breath Volume	1.347	22:54	
Air Blank	0.000	22:54	
Air Blank	0.000	22:56	
Subject Sample	# 2	22:56	Subject Test Refused
Breath Volume	0.000	22:56	
Air Blank	0.000	22:57	
Control Test	0.079	22:57	
Air Blank	0.000	22:57	
Diagnostics Check	OK	22:58	

Data Download Date/Time: 09/29/2018 12:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRINGTON,ALEX,M
BREATH TEST OPERATOR AGENCY: Belleview PD
ARRESTING OFFICER: HARRINGTON,ALEX
ARRESTING OFFICER AGENCY: Belleview PD
LAST AGENCY INSPECTION: 08/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: DEWAR

DATE/TIME OF TEST: 09/08/2018 03:26:23

Diagnostics Check	OK	03:20
Air Blank	0.000	03:20
Control Test	0.080	03:20
Air Blank	0.000	03:21
Subject Sample # 1	0.078	03:21
Breath Volume	3.554	03:21
Air Blank	0.000	03:22
Air Blank	0.000	03:24
Subject Sample # 2	0.077	03:24
Breath Volume	3.523	03:24
Air Blank	0.000	03:25
Control Test	0.079	03:25
Air Blank	0.000	03:26
Diagnostics Check	OK	03:26

Data Download Date/Time: 09/29/2018 12:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LACOGNATA,CHRISTINA,T
BREATH TEST OPERATOR AGENCY: Belleview PD
ARRESTING OFFICER: LACOGNATA,CHRISTINA
ARRESTING OFFICER AGENCY: Belleview PD
LAST AGENCY INSPECTION: 08/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: GONZALEZ-SANCHEZ

DATE/TIME OF TEST: 09/08/2018 04:41:42

Diagnostics Check	OK	04:30	
Air Blank	0.000	04:30	
Control Test	0.080	04:31	
Air Blank	0.000	04:31	
Subject Sample	# 1	04:34	No Sample Provided
Breath Volume	0.000	04:34	
Air Blank	0.000	04:35	
Air Blank	0.000	04:37	
Subject Sample	# 2	04:40	No Sample Provided
Breath Volume	0.000	04:40	
Air Blank	0.000	04:40	
Control Test	0.080	04:41	
Air Blank	0.000	04:41	
Diagnostics Check	OK	04:41	

Data Download Date/Time: 09/29/2018 12:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTON,TIMOTHY,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MARTON,TIMOTHY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: PERRY

DATE/TIME OF TEST: 09/08/2018 05:07:30

Diagnostics Check	OK	05:04	
Air Blank	0.000	05:05	
Control Test	0.080	05:05	
Air Blank	0.000	05:05	
Subject Sample	# 1	05:06	Subject Test Refused
Breath Volume	0.000	05:06	
Air Blank	0.000	05:06	
Control Test	0.079	05:06	
Air Blank	0.000	05:07	
Diagnostics Check	OK	05:07	

Data Download Date/Time: 09/29/2018 12:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPE,CALVIN,W
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: POPE,CALVIN
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 03:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: SANTIAGO-CRUZ

DATE/TIME OF TEST: 09/09/2018 05:09:50

Diagnostics Check	OK	05:02
Air Blank	0.000	05:02
Control Test	0.080	05:03
Air Blank	0.000	05:03
Subject Sample # 1	0.128	05:04
Breath Volume	2.335	05:04
Air Blank	0.000	05:05
Air Blank	0.000	05:07
Subject Sample # 2	0.128	05:08
Breath Volume	1.109	05:08
Air Blank	0.000	05:08
Control Test	0.079	05:09
Air Blank	0.000	05:09
Diagnostics Check	OK	05:09

Data Download Date/Time: 09/29/2018 12:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRINGTON,ALEX,M
BREATH TEST OPERATOR AGENCY: Belleview PD
ARRESTING OFFICER: HARRINGTON,ALEX
ARRESTING OFFICER AGENCY: Belleview PD
LAST AGENCY INSPECTION: 08/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: HARBUCK

DATE/TIME OF TEST: 09/10/2018 02:27:20

Diagnostics Check	OK	02:20
Air Blank	0.000	02:20
Control Test	0.080	02:21
Air Blank	0.000	02:21
Subject Sample # 1	0.010	02:23
Breath Volume	1.386	02:23
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample # 2	0.008	02:25
Breath Volume	1.148	02:25
Air Blank	0.000	02:26
Control Test	0.079	02:26
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

Data Download Date/Time: 09/29/2018 12:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MERCER,ISAAC,V
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MERCER,ISAAC
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 09/11/2018 03:52:56

Diagnostics Check	OK	03:41	
Air Blank	0.000	03:42	
Control Test	0.081	03:42	
Air Blank	0.000	03:43	
Subject Sample # 1	0.148	03:46	Volume Not Met
Breath Volume	0.031	03:46	
Air Blank	0.000	03:47	
Air Blank	0.000	03:48	
Subject Sample # 2	0.175	03:51	
Breath Volume	2.230	03:51	
Air Blank		03:52	Purge Fail
Air Blank	0.000	03:52	

Data Download Date/Time: 09/29/2018 12:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MERCER,ISAAC,V
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MERCER,ISAAC
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 09/11/2018 04:20:42

Diagnostics Check	OK	04:14
Air Blank	0.000	04:14
Control Test	0.077	04:15
Air Blank	0.000	04:15
Subject Sample # 1	0.194	04:16
Breath Volume	1.273	04:16
Air Blank	0.000	04:16
Air Blank	0.000	04:18
Subject Sample # 2	0.180	04:19
Breath Volume	1.605	04:19
Air Blank	0.000	04:19
Control Test	0.078	04:20
Air Blank	0.000	04:20
Diagnostics Check	OK	04:20

Data Download Date/Time: 09/29/2018 12:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOLLAZZO,C,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: DEL RUSSO,TROOPER
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 21:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:43

SUBJECT'S LAST NAME: SLAVIK

DATE/TIME OF TEST: 09/12/2018 23:12:50

Diagnostics Check	OK	23:06
Air Blank	0.000	23:07
Control Test	0.081	23:07
Air Blank	0.000	23:08
Subject Sample # 1	0.000	23:08
Breath Volume	2.160	23:08
Air Blank	0.000	23:09
Air Blank	0.000	23:11
Subject Sample # 2	0.000	23:11
Breath Volume	1.972	23:11
Air Blank	0.000	23:11
Control Test	0.080	23:12
Air Blank	0.000	23:12
Diagnostics Check	OK	23:12

Data Download Date/Time: 09/29/2018 12:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRINGTON,ALEX,M
BREATH TEST OPERATOR AGENCY: Belleview PD
ARRESTING OFFICER: HARRINGTON,ALEX
ARRESTING OFFICER AGENCY: Belleview PD
LAST AGENCY INSPECTION: 08/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: RAMSEY

DATE/TIME OF TEST: 09/13/2018 02:57:24

Diagnostics Check	OK	02:51
Air Blank	0.000	02:51
Control Test	0.080	02:51
Air Blank	0.000	02:52
Subject Sample # 1	0.128	02:52
Breath Volume	1.148	02:52
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample # 2	0.121	02:55
Breath Volume	1.175	02:55
Air Blank	0.000	02:56
Control Test	0.080	02:56
Air Blank	0.000	02:57
Diagnostics Check	OK	02:57

Data Download Date/Time: 09/29/2018 12:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GENTHER, EMILY, S
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: CABRALES, ERIK
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 08/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 18:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:04

SUBJECT'S LAST NAME: WITTMAN

DATE/TIME OF TEST: 09/14/2018 19:33:51

Diagnostics Check	OK	19:27
Air Blank	0.000	19:28
Control Test	0.080	19:28
Air Blank	0.000	19:28
Subject Sample # 1	0.369	19:29
Breath Volume	2.687	19:29
Air Blank	0.000	19:29
Air Blank	0.000	19:31
Subject Sample # 2	0.366	19:32
Breath Volume	2.425	19:32
Air Blank	0.000	19:32
Control Test	0.079	19:33
Air Blank	0.000	19:33
Diagnostics Check	OK	19:33

Data Download Date/Time: 09/29/2018 12:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EARNEST, WILLIAM, M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: EARNEST, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: MCLEROY

DATE/TIME OF TEST: 09/16/2018 00:57:00

Diagnostics Check	OK	00:49
Air Blank	0.000	00:50
Control Test	0.081	00:50
Air Blank	0.000	00:51
Subject Sample # 1	0.142	00:51
Breath Volume	1.113	00:51
Air Blank	0.000	00:52
Air Blank	0.000	00:54
Subject Sample # 2	0.141	00:55
Breath Volume	1.789	00:55
Air Blank	0.000	00:56
Control Test	0.080	00:56
Air Blank	0.000	00:56
Diagnostics Check	OK	00:57

Data Download Date/Time: 09/29/2018 12:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,MICAH,
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: DIAZ,JOSEPH
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 09:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:10

SUBJECT'S LAST NAME: CROWELL

DATE/TIME OF TEST: 09/16/2018 10:36:56

Diagnostics Check	OK	10:31
Air Blank	0.000	10:31
Control Test	0.080	10:31
Air Blank	0.000	10:32
Subject Sample # 1	0.000	10:32
Breath Volume	1.222	10:32
Air Blank	0.000	10:33
Air Blank	0.000	10:35
Subject Sample # 2	0.000	10:35
Breath Volume	1.218	10:35
Air Blank	0.000	10:36
Control Test	0.082	10:36
Air Blank	0.000	10:36
Diagnostics Check	OK	10:36

Data Download Date/Time: 09/29/2018 12:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPE,CALVIN,W
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: ESQUIVEL,ANTHONY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 20:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:11

SUBJECT'S LAST NAME: HALL JR

DATE/TIME OF TEST: 09/17/2018 21:38:33

Diagnostics Check	OK	21:32
Air Blank	0.000	21:32
Control Test	0.081	21:32
Air Blank	0.000	21:33
Subject Sample # 1	0.128	21:33
Breath Volume	3.066	21:33
Air Blank	0.000	21:34
Air Blank	0.000	21:36
Subject Sample # 2	0.131	21:36
Breath Volume	2.898	21:36
Air Blank	0.000	21:37
Control Test	0.080	21:37
Air Blank	0.000	21:38
Diagnostics Check	OK	21:38

Data Download Date/Time: 09/29/2018 12:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOLLAZZO,C,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SOLLAZZO,C
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: CARSON

DATE/TIME OF TEST: 09/18/2018 03:28:27

Diagnostics Check	OK	03:22
Air Blank	0.000	03:22
Control Test	0.081	03:23
Air Blank	0.000	03:23
Subject Sample # 1	0.160	03:24
Breath Volume	1.730	03:24
Air Blank	0.000	03:24
Air Blank	0.000	03:26
Subject Sample # 2	0.170	03:26
Breath Volume	1.890	03:26
Air Blank	0.000	03:27
Control Test	0.080	03:27
Air Blank	0.000	03:28
Diagnostics Check	OK	03:28

Data Download Date/Time: 09/29/2018 12:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BALL,CODY,W
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: EARNEST,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: MCCLAIN

DATE/TIME OF TEST: 09/20/2018 00:44:57

Diagnostics Check	OK	00:35
Air Blank	0.000	00:35
Control Test	0.079	00:36
Air Blank	0.000	00:36
Subject Sample # 1	0.227	00:39
Breath Volume	3.441	00:39
Air Blank	0.000	00:40
Air Blank	0.000	00:41
Subject Sample # 2	0.224	00:43
Breath Volume	2.976	00:43
Air Blank	0.000	00:44
Control Test	0.079	00:44
Air Blank	0.000	00:44
Diagnostics Check	OK	00:44

Data Download Date/Time: 09/29/2018 12:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BALL,CODY,W
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BALL,CODY
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: COPELAND

DATE/TIME OF TEST: 09/20/2018 02:26:02

Diagnostics Check	OK	02:22	
Air Blank	0.000	02:23	
Control Test	0.080	02:23	
Air Blank	0.000	02:23	
Subject Sample	# 1	02:24	Subject Test Refused
Breath Volume	0.000	02:24	
Air Blank	0.000	02:25	
Control Test	0.081	02:25	
Air Blank	0.000	02:25	
Diagnostics Check	OK	02:26	

Data Download Date/Time: 09/29/2018 12:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REID JR,MICHAEL,G
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: REID JR,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 17:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: SETSMA

DATE/TIME OF TEST: 09/21/2018 19:48:10

Diagnostics Check	OK	19:42
Air Blank	0.000	19:42
Control Test	0.081	19:42
Air Blank	0.000	19:43
Subject Sample # 1	0.179	19:43
Breath Volume	2.523	19:43
Air Blank	0.000	19:44
Air Blank	0.000	19:46
Subject Sample # 2	0.175	19:46
Breath Volume	2.605	19:46
Air Blank	0.000	19:47
Control Test	0.081	19:47
Air Blank	0.000	19:48
Diagnostics Check	OK	19:48

Data Download Date/Time: 09/29/2018 12:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOLLAZZO,C,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SOLLAZZO,C
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 09/22/2018 00:41:24

Diagnostics Check	OK	00:38	
Air Blank	0.000	00:39	
Control Test	0.081	00:39	
Air Blank	0.000	00:39	
Subject Sample	# 1	00:40	Subject Test Refused
Breath Volume	0.000	00:40	
Air Blank	0.000	00:40	
Control Test	0.081	00:40	
Air Blank	0.000	00:41	
Diagnostics Check	OK	00:41	

Data Download Date/Time: 09/29/2018 12:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BALL,CODY,W
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: LANDER,JAMES
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 11:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:25

SUBJECT'S LAST NAME: MCKNIGHT

DATE/TIME OF TEST: 09/22/2018 16:06:44

Diagnostics Check	OK	16:00
Air Blank	0.000	16:00
Control Test	0.081	16:01
Air Blank	0.000	16:01
Subject Sample # 1	0.000	16:02
Breath Volume	1.546	16:02
Air Blank	0.000	16:02
Air Blank	0.000	16:04
Subject Sample # 2	0.000	16:05
Breath Volume	1.410	16:05
Air Blank	0.000	16:05
Control Test	0.081	16:06
Air Blank	0.000	16:06
Diagnostics Check	OK	16:06

Data Download Date/Time: 09/29/2018 12:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOLLAZZO,C,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SOLLAZZO,C
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 20:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:41

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 09/22/2018 22:05:56

Diagnostics Check	OK	22:03	
Air Blank	0.000	22:03	
Control Test	0.081	22:03	
Air Blank	0.000	22:04	
Subject Sample	# 1	22:04	Subject Test Refused
Breath Volume	0.000	22:04	
Air Blank	0.000	22:05	
Control Test	0.080	22:05	
Air Blank	0.000	22:05	
Diagnostics Check	OK	22:05	

Data Download Date/Time: 09/29/2018 12:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAFFERTY,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: RAFFERTY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 09/26/2018 02:05:38

Diagnostics Check	OK	01:59
Air Blank	0.000	01:59
Control Test	0.081	02:00
Air Blank	0.000	02:00
Subject Sample # 1	0.138	02:01
Breath Volume	1.343	02:01
Air Blank	0.000	02:01
Air Blank	0.000	02:03
Subject Sample # 2	0.142	02:03
Breath Volume	2.054	02:03
Air Blank	0.000	02:04
Control Test	0.079	02:05
Air Blank	0.000	02:05
Diagnostics Check	OK	02:05

Data Download Date/Time: 09/29/2018 12:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPE,CALVIN,W
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: POPE,CALVIN
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 18:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:34

SUBJECT'S LAST NAME: BICKHAUS

DATE/TIME OF TEST: 09/26/2018 19:57:08

Diagnostics Check	OK	19:54	
Air Blank	0.000	19:54	
Control Test	0.081	19:55	
Air Blank	0.000	19:55	
Subject Sample	# 1	19:55	Subject Test Refused
Breath Volume	0.000	19:55	
Air Blank	0.000	19:56	
Control Test	0.080	19:56	
Air Blank	0.000	19:57	
Diagnostics Check	OK	19:57	

Data Download Date/Time: 09/29/2018 12:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOLLAZZO,C,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SOLLAZZO,C
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: MCCOMBS

DATE/TIME OF TEST: 09/27/2018 23:09:37

Diagnostics Check	OK	23:03
Air Blank	0.000	23:04
Control Test	0.081	23:04
Air Blank	0.000	23:04
Subject Sample # 1	0.000	23:05
Breath Volume	1.496	23:05
Air Blank	0.000	23:05
Air Blank	0.000	23:07
Subject Sample # 2	0.000	23:08
Breath Volume	1.972	23:08
Air Blank	0.000	23:08
Control Test	0.081	23:09
Air Blank	0.000	23:09
Diagnostics Check	OK	23:09

Data Download Date/Time: 09/29/2018 12:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LETSON,REBECCA,G
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: LETSON,REBECCA
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 08/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: SYKES

DATE/TIME OF TEST: 09/29/2018 03:57:14

Diagnostics Check	OK	03:51
Air Blank	0.000	03:51
Control Test	0.081	03:51
Air Blank	0.000	03:52
Subject Sample # 1	0.229	03:52
Breath Volume	3.847	03:52
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample # 2	0.215	03:55
Breath Volume	3.199	03:55
Air Blank	0.000	03:56
Control Test	0.080	03:56
Air Blank	0.000	03:57
Diagnostics Check	OK	03:57

Data Download Date/Time: 09/29/2018 12:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LETSON,REBECCA,G
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: LETSON,REBECCA
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 09/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: SHELTON

DATE/TIME OF TEST: 09/30/2018 03:33:21

Diagnostics Check	OK	03:27
Air Blank	0.000	03:27
Control Test	0.081	03:27
Air Blank	0.000	03:28
Subject Sample # 1	0.097	03:28
Breath Volume	2.246	03:28
Air Blank	0.000	03:29
Air Blank	0.000	03:31
Subject Sample # 2	0.093	03:31
Breath Volume	2.097	03:31
Air Blank	0.000	03:32
Control Test	0.080	03:32
Air Blank	0.000	03:33
Diagnostics Check	OK	03:33

Data Download Date/Time: 10/30/2018 16:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LETSON,REBECCA,G
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: LETSON,REBECCA
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 09/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: MARKHAM

DATE/TIME OF TEST: 09/30/2018 23:58:34

Diagnostics Check	OK	23:52
Air Blank	0.000	23:52
Control Test	0.080	23:53
Air Blank	0.000	23:53
Subject Sample # 1	0.192	23:54
Breath Volume	3.152	23:54
Air Blank	0.000	23:54
Air Blank	0.000	23:56
Subject Sample # 2	0.191	23:57
Breath Volume	3.015	23:57
Air Blank	0.000	23:57
Control Test	0.079	23:57
Air Blank	0.000	23:58
Diagnostics Check	OK	23:58

Data Download Date/Time: 10/30/2018 16:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LYON, M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 09/07/2018 05:15:19

Diagnostics Check	OK	05:05	
Air Blank	0.000	05:05	
Control Test	0.079	05:05	
Air Blank	0.000	05:06	
Subject Sample	# 1	05:09	No Sample Provided
Breath Volume	0.000	05:09	
Air Blank	0.000	05:10	
Air Blank	0.000	05:12	
Subject Sample	# 2	05:13	Subject Test Refused
Breath Volume	0.000	05:13	
Air Blank	0.000	05:14	
Control Test	0.079	05:14	
Air Blank	0.000	05:15	
Diagnostics Check	OK	05:15	

Data Download Date/Time: 10/04/2018 11:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DRISCOLL,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: WILCOX

DATE/TIME OF TEST: 09/07/2018 21:37:36

Diagnostics Check	OK	21:29
Air Blank	0.000	21:30
Control Test	0.079	21:30
Air Blank	0.000	21:31
Subject Sample # 1	0.000	21:32
Breath Volume	2.507	21:32
Air Blank	0.000	21:33
Air Blank	0.000	21:35
Subject Sample # 2	0.000	21:36
Breath Volume	2.496	21:36
Air Blank	0.000	21:36
Control Test	0.078	21:37
Air Blank	0.000	21:37
Diagnostics Check	OK	21:37

Data Download Date/Time: 10/04/2018 11:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE, S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: CANCEL

DATE/TIME OF TEST: 09/08/2018 01:19:51

Diagnostics Check	OK	01:13
Air Blank	0.000	01:13
Control Test	0.079	01:14
Air Blank	0.000	01:14
Subject Sample # 1	0.172	01:15
Breath Volume	2.281	01:15
Air Blank	0.000	01:15
Air Blank	0.000	01:17
Subject Sample # 2	0.177	01:18
Breath Volume	2.164	01:18
Air Blank	0.000	01:18
Control Test	0.079	01:19
Air Blank	0.000	01:19
Diagnostics Check	OK	01:19

Data Download Date/Time: 10/04/2018 11:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MOREJON,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: TOMER

DATE/TIME OF TEST: 09/09/2018 02:18:35

Diagnostics Check	OK	02:11
Air Blank	0.000	02:11
Control Test	0.080	02:12
Air Blank	0.000	02:12
Subject Sample # 1	0.158	02:13
Breath Volume	2.859	02:13
Air Blank	0.000	02:14
Air Blank	0.000	02:15
Subject Sample # 2	0.145	02:17
Breath Volume	1.816	02:17
Air Blank	0.000	02:17
Control Test	0.079	02:18
Air Blank	0.000	02:18
Diagnostics Check	OK	02:18

Data Download Date/Time: 10/04/2018 11:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: HARRIS WILLINGHAM

DATE/TIME OF TEST: 09/09/2018 04:55:16

Diagnostics Check	OK	04:49
Air Blank	0.000	04:49
Control Test	0.079	04:49
Air Blank	0.000	04:50
Subject Sample # 1	0.129	04:50
Breath Volume	1.359	04:50
Air Blank	0.000	04:51
Air Blank	0.000	04:53
Subject Sample # 2	0.124	04:53
Breath Volume	1.480	04:53
Air Blank	0.000	04:54
Control Test	0.078	04:54
Air Blank	0.000	04:55
Diagnostics Check	OK	04:55

Data Download Date/Time: 10/04/2018 11:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIMPKINS,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:46

SUBJECT'S LAST NAME: LOPEZ-CUEVAS

DATE/TIME OF TEST: 09/09/2018 05:15:31

Diagnostics Check	OK	05:07
Air Blank	0.000	05:07
Control Test	0.079	05:08
Air Blank	0.000	05:08
Subject Sample # 1	0.237	05:10
Breath Volume	1.761	05:10
Air Blank	0.000	05:11
Air Blank	0.000	05:13
Subject Sample # 2	0.226	05:14
Breath Volume	2.355	05:14
Air Blank	0.000	05:14
Control Test	0.077	05:14
Air Blank	0.000	05:15
Diagnostics Check	OK	05:15

Data Download Date/Time: 10/04/2018 11:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NA,NA
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: ADMIN
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:30

SUBJECT'S LAST NAME: OQUENDO

DATE/TIME OF TEST: 09/09/2018 09:19:44

Diagnostics Check	OK	09:13
Air Blank	0.000	09:14
Control Test	0.079	09:14
Air Blank	0.000	09:14
Subject Sample # 1	0.043	09:15
Breath Volume	2.527	09:15
Air Blank	0.000	09:15
Air Blank	0.000	09:17
Subject Sample # 2	0.041	09:18
Breath Volume	2.484	09:18
Air Blank	0.000	09:18
Control Test	0.079	09:19
Air Blank	0.000	09:19
Diagnostics Check	OK	09:19

Data Download Date/Time: 10/04/2018 11:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, REBECCA, J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HORNE, A
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:08

SUBJECT'S LAST NAME: ROSSON

DATE/TIME OF TEST: 09/09/2018 17:36:56

Diagnostics Check	OK	17:28	
Air Blank	0.000	17:29	
Control Test	0.080	17:29	
Air Blank	0.000	17:29	
Subject Sample	# 1	17:32	No Sample Provided
Breath Volume	0.000	17:32	
Air Blank	0.000	17:33	
Air Blank	0.000	17:35	
Subject Sample	# 2	17:35	Subject Test Refused
Breath Volume	0.000	17:35	
Air Blank	0.000	17:36	
Control Test	0.080	17:36	
Air Blank	0.000	17:36	
Diagnostics Check	OK	17:36	

Data Download Date/Time: 10/04/2018 11:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SULLIMAN,P
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: TEJADA

DATE/TIME OF TEST: 09/09/2018 19:58:17

Diagnostics Check	OK	19:51
Air Blank	0.000	19:52
Control Test	0.081	19:52
Air Blank	0.000	19:52
Subject Sample # 1	0.169	19:53
Breath Volume	1.246	19:53
Air Blank	0.000	19:54
Air Blank	0.000	19:56
Subject Sample # 2	0.167	19:56
Breath Volume	2.675	19:56
Air Blank	0.000	19:57
Control Test	0.079	19:57
Air Blank	0.000	19:58
Diagnostics Check	OK	19:58

Data Download Date/Time: 10/04/2018 11:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIMPKINS,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: NUNEZ

DATE/TIME OF TEST: 09/09/2018 21:17:07

Diagnostics Check	OK	21:10
Air Blank	0.000	21:10
Control Test	0.079	21:10
Air Blank	0.000	21:11
Subject Sample # 1	0.232	21:12
Breath Volume	1.339	21:12
Air Blank	0.000	21:13
Air Blank	0.000	21:15
Subject Sample # 2	0.227	21:15
Breath Volume	1.816	21:15
Air Blank	0.000	21:16
Control Test	0.079	21:16
Air Blank	0.000	21:17
Diagnostics Check	OK	21:17

Data Download Date/Time: 10/04/2018 11:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:51

SUBJECT'S LAST NAME: SUAREZ

DATE/TIME OF TEST: 09/11/2018 19:19:33

Diagnostics Check	OK	19:11
Air Blank	0.000	19:11
Control Test	0.079	19:12
Air Blank	0.000	19:12
Subject Sample # 1	0.271	19:14
Breath Volume	2.203	19:14
Air Blank	0.000	19:15
Air Blank	0.000	19:17
Subject Sample # 2	0.265	19:18
Breath Volume	2.398	19:18
Air Blank	0.000	19:18
Control Test	0.080	19:18
Air Blank	0.000	19:19
Diagnostics Check	OK	19:19

Data Download Date/Time: 10/04/2018 11:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SCHILLER,E
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 09/11/2018 23:49:56

Diagnostics Check	OK	23:43
Air Blank	0.000	23:43
Control Test	0.080	23:44
Air Blank	0.000	23:44
Subject Sample # 1	0.000	23:45
Breath Volume	1.421	23:45
Air Blank	0.000	23:45
Air Blank	0.000	23:47
Subject Sample # 2	0.000	23:48
Breath Volume	1.320	23:48
Air Blank	0.000	23:49
Control Test	0.079	23:49
Air Blank	0.000	23:49
Diagnostics Check	OK	23:49

Data Download Date/Time: 10/04/2018 11:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DRISCOLL, B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: BOWEN

DATE/TIME OF TEST: 09/13/2018 02:50:53

Diagnostics Check	OK	02:44
Air Blank	0.000	02:44
Control Test	0.079	02:45
Air Blank	0.000	02:45
Subject Sample # 1	0.037	02:46
Breath Volume	2.062	02:46
Air Blank	0.000	02:46
Air Blank	0.000	02:48
Subject Sample # 2	0.037	02:49
Breath Volume	2.480	02:49
Air Blank	0.000	02:49
Control Test	0.079	02:50
Air Blank	0.000	02:50
Diagnostics Check	OK	02:50

Data Download Date/Time: 10/04/2018 11:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRIZENDINE,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: HERDOCIA

DATE/TIME OF TEST: 09/14/2018 23:05:21

Diagnostics Check	OK	22:58
Air Blank	0.000	22:58
Control Test	0.080	22:59
Air Blank	0.000	22:59
Subject Sample # 1	0.136	23:00
Breath Volume	1.667	23:00
Air Blank	0.000	23:01
Air Blank	0.000	23:02
Subject Sample # 2	0.139	23:03
Breath Volume	1.894	23:03
Air Blank	0.000	23:04
Control Test	0.079	23:04
Air Blank	0.000	23:05
Diagnostics Check	OK	23:05

Data Download Date/Time: 10/04/2018 11:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NAVE,R
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 09/15/2018 03:32:12

Diagnostics Check	OK	03:25
Air Blank	0.000	03:25
Control Test	0.079	03:26
Air Blank	0.000	03:26
Subject Sample # 1	0.164	03:27
Breath Volume	1.632	03:27
Air Blank	0.000	03:28
Air Blank	0.000	03:30
Subject Sample # 2	0.164	03:30
Breath Volume	1.699	03:30
Air Blank	0.000	03:31
Control Test	0.079	03:31
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

Data Download Date/Time: 10/04/2018 11:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: CARRASQUILLO

DATE/TIME OF TEST: 09/15/2018 05:17:16

Diagnostics Check	OK	05:10
Air Blank	0.000	05:11
Control Test	0.080	05:11
Air Blank	0.000	05:12
Subject Sample # 1	0.236	05:12
Breath Volume	2.632	05:12
Air Blank	0.000	05:13
Air Blank	0.000	05:15
Subject Sample # 2	0.237	05:15
Breath Volume	3.378	05:15
Air Blank	0.000	05:16
Control Test	0.079	05:16
Air Blank	0.000	05:17
Diagnostics Check	OK	05:17

Data Download Date/Time: 10/04/2018 11:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,ANGELA,V
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,DAVID
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 06:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:59

SUBJECT'S LAST NAME: BOYINGTON

DATE/TIME OF TEST: 09/15/2018 07:26:52

Diagnostics Check	OK	07:20
Air Blank	0.000	07:20
Control Test	0.079	07:21
Air Blank	0.000	07:21
Subject Sample # 1	0.000	07:22
Breath Volume	2.882	07:22
Air Blank	0.000	07:23
Air Blank	0.000	07:25
Subject Sample # 2	0.000	07:25
Breath Volume	2.765	07:25
Air Blank	0.000	07:25
Control Test	0.079	07:26
Air Blank	0.000	07:26
Diagnostics Check	OK	07:26

Data Download Date/Time: 10/04/2018 11:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: STEAD, K
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:41

SUBJECT'S LAST NAME: MILEY

DATE/TIME OF TEST: 09/15/2018 19:09:56

Diagnostics Check	OK	19:02
Air Blank	0.000	19:03
Control Test	0.080	19:03
Air Blank	0.000	19:04
Subject Sample # 1	0.000	19:04
Breath Volume	1.839	19:04
Air Blank	0.000	19:05
Air Blank	0.000	19:07
Subject Sample # 2	0.000	19:08
Breath Volume	1.945	19:08
Air Blank	0.000	19:09
Control Test	0.080	19:09
Air Blank	0.000	19:09
Diagnostics Check	OK	19:09

Data Download Date/Time: 10/04/2018 11:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:59

SUBJECT'S LAST NAME: IRMLER

DATE/TIME OF TEST: 09/15/2018 22:27:51

Diagnostics Check	OK	22:20
Air Blank	0.000	22:20
Control Test	0.080	22:20
Air Blank	0.000	22:21
Subject Sample # 1	0.000	22:22
Breath Volume	1.449	22:22
Air Blank	0.000	22:23
Air Blank	0.000	22:25
Subject Sample # 2	0.000	22:26
Breath Volume	1.421	22:26
Air Blank	0.000	22:26
Control Test	0.079	22:27
Air Blank	0.000	22:27
Diagnostics Check	OK	22:27

Data Download Date/Time: 10/04/2018 11:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:33

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 09/16/2018 05:11:01

Diagnostics Check	OK	05:01
Air Blank	0.000	05:02
Control Test	0.079	05:02
Air Blank	0.000	05:03
Subject Sample # 1	0.164	05:05
Breath Volume	1.210	05:05
Air Blank	0.000	05:06
Air Blank	0.000	05:08
Subject Sample # 2	0.169	05:09
Breath Volume	1.679	05:09
Air Blank	0.000	05:10
Control Test	0.078	05:10
Air Blank	0.000	05:10
Diagnostics Check	OK	05:11

Data Download Date/Time: 10/04/2018 11:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY, D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: GARRETT JR

DATE/TIME OF TEST: 09/16/2018 05:32:34

Diagnostics Check	OK	05:25
Air Blank	0.000	05:26
Control Test	0.079	05:26
Air Blank	0.000	05:27
Subject Sample # 1	0.147	05:28
Breath Volume	4.625	05:28
Air Blank	0.000	05:28
Air Blank	0.000	05:30
Subject Sample # 2	0.146	05:31
Breath Volume	3.414	05:31
Air Blank	0.000	05:31
Control Test	0.078	05:31
Air Blank	0.000	05:32
Diagnostics Check	OK	05:32

Data Download Date/Time: 10/04/2018 11:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,BARBARA,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: STEAD,K
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:45

SUBJECT'S LAST NAME: WIGGINS

DATE/TIME OF TEST: 09/16/2018 16:19:43

Diagnostics Check	OK	16:11
Air Blank	0.000	16:11
Control Test	0.080	16:11
Air Blank	0.000	16:12
Subject Sample # 1	0.168	16:14
Breath Volume	1.644	16:14
Air Blank	0.000	16:15
Air Blank	0.000	16:17
Subject Sample # 2	0.178	16:18
Breath Volume	2.273	16:18
Air Blank	0.000	16:18
Control Test	0.079	16:19
Air Blank	0.000	16:19
Diagnostics Check	OK	16:19

Data Download Date/Time: 10/04/2018 11:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: SANTOS

DATE/TIME OF TEST: 09/17/2018 01:18:07

Diagnostics Check	OK	01:11
Air Blank	0.000	01:11
Control Test	0.079	01:12
Air Blank	0.000	01:12
Subject Sample # 1	0.000	01:13
Breath Volume	3.148	01:13
Air Blank	0.000	01:14
Air Blank	0.000	01:16
Subject Sample # 2	0.000	01:16
Breath Volume	3.128	01:16
Air Blank	0.000	01:17
Control Test	0.079	01:17
Air Blank	0.000	01:18
Diagnostics Check	OK	01:18

Data Download Date/Time: 10/04/2018 11:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: COSTA,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: HARE

DATE/TIME OF TEST: 09/17/2018 23:30:44

Diagnostics Check	OK	23:26	
Air Blank	0.000	23:27	
Control Test	0.079	23:27	
Air Blank	0.000	23:28	
Subject Sample	# 1	23:29	Subject Test Refused
Breath Volume	0.000	23:29	
Air Blank	0.000	23:29	
Control Test	0.080	23:30	
Air Blank	0.000	23:30	
Diagnostics Check	OK	23:30	

Data Download Date/Time: 10/04/2018 11:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 09/18/2018 05:25:48

Diagnostics Check	OK	05:15
Air Blank	0.000	05:16
Control Test	0.079	05:16
Air Blank	0.000	05:17
Subject Sample # 1	0.158	05:20
Breath Volume	2.179	05:20
Air Blank	0.000	05:20
Air Blank	0.000	05:22
Subject Sample # 2	0.151	05:24
Breath Volume	1.703	05:24
Air Blank	0.000	05:24
Control Test	0.078	05:25
Air Blank	0.000	05:25
Diagnostics Check	OK	05:25

Data Download Date/Time: 10/04/2018 11:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANDERSON,C
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:43

SUBJECT'S LAST NAME: BECERRA DEOLARTE

DATE/TIME OF TEST: 09/20/2018 23:09:22

Diagnostics Check	OK	23:03
Air Blank	0.000	23:03
Control Test	0.079	23:03
Air Blank	0.000	23:04
Subject Sample # 1	0.242	23:04
Breath Volume	2.742	23:04
Air Blank	0.000	23:05
Air Blank	0.000	23:07
Subject Sample # 2	0.240	23:07
Breath Volume	3.125	23:07
Air Blank	0.000	23:08
Control Test	0.079	23:08
Air Blank	0.000	23:09
Diagnostics Check	OK	23:09

Data Download Date/Time: 10/04/2018 11:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: ALTURKI

DATE/TIME OF TEST: 09/21/2018 03:41:45

Diagnostics Check	OK	03:37	
Air Blank	0.000	03:38	
Control Test	0.080	03:38	
Air Blank	0.000	03:38	
Subject Sample	# 1	03:40	Subject Test Refused
Breath Volume	0.000	03:40	
Air Blank	0.000	03:40	
Control Test	0.079	03:41	
Air Blank	0.000	03:41	
Diagnostics Check	OK	03:41	

Data Download Date/Time: 10/04/2018 11:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MOREJON, A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 09/21/2018 22:23:04

Diagnostics Check	OK	22:16
Air Blank	0.000	22:17
Control Test	0.079	22:17
Air Blank	0.000	22:17
Subject Sample # 1	0.211	22:18
Breath Volume	2.601	22:18
Air Blank	0.000	22:18
Air Blank	0.000	22:20
Subject Sample # 2	0.220	22:21
Breath Volume	3.476	22:21
Air Blank	0.000	22:22
Control Test	0.079	22:22
Air Blank	0.000	22:22
Diagnostics Check	OK	22:23

Data Download Date/Time: 10/04/2018 11:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DRISCOLL,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: RUSSO

DATE/TIME OF TEST: 09/22/2018 05:00:00

Diagnostics Check	OK	04:56	
Air Blank	0.000	04:56	
Control Test	0.079	04:57	
Air Blank	0.000	04:57	
Subject Sample	# 1	04:58	Subject Test Refused
Breath Volume	0.000	04:58	
Air Blank	0.000	04:59	
Control Test	0.080	04:59	
Air Blank	0.000	04:59	
Diagnostics Check	OK	05:00	

Data Download Date/Time: 10/04/2018 11:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANDERSON,C
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: PAZ-CRUZ

DATE/TIME OF TEST: 09/22/2018 22:17:59

Diagnostics Check	OK	22:09
Air Blank	0.000	22:09
Control Test	0.080	22:09
Air Blank	0.000	22:10
Subject Sample # 1	0.184	22:12
Breath Volume	1.609	22:12
Air Blank	0.000	22:13
Air Blank	0.000	22:15
Subject Sample # 2	0.183	22:16
Breath Volume	1.273	22:16
Air Blank	0.000	22:17
Control Test	0.078	22:17
Air Blank	0.000	22:17
Diagnostics Check	OK	22:17

Data Download Date/Time: 10/04/2018 11:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DRISCOLL,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: GOODING

DATE/TIME OF TEST: 09/22/2018 23:46:58

Diagnostics Check	OK	23:39
Air Blank	0.000	23:40
Control Test	0.080	23:40
Air Blank	0.000	23:41
Subject Sample # 1	0.224	23:42
Breath Volume	1.367	23:42
Air Blank	0.000	23:42
Air Blank	0.000	23:44
Subject Sample # 2	0.228	23:45
Breath Volume	2.125	23:45
Air Blank	0.000	23:46
Control Test	0.078	23:46
Air Blank	0.000	23:46
Diagnostics Check	OK	23:46

Data Download Date/Time: 10/04/2018 11:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIMPKINS,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: GUILLEN

DATE/TIME OF TEST: 09/23/2018 02:42:58

Diagnostics Check	OK	02:33
Air Blank	0.000	02:33
Control Test	0.079	02:34
Air Blank	0.000	02:34
Subject Sample # 1	0.106	02:37
Breath Volume	1.402	02:37
Air Blank	0.000	02:37
Air Blank	0.000	02:39
Subject Sample # 2	0.114	02:41
Breath Volume	1.851	02:41
Air Blank	0.000	02:42
Control Test	0.079	02:42
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 10/04/2018 11:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SULAIMAN,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:51

SUBJECT'S LAST NAME: LAINEZ-ALVARADO

DATE/TIME OF TEST: 09/23/2018 05:20:47

Diagnostics Check	OK	05:11
Air Blank	0.000	05:12
Control Test	0.080	05:12
Air Blank	0.000	05:13
Subject Sample # 1	0.163	05:15
Breath Volume	1.582	05:15
Air Blank	0.000	05:16
Air Blank	0.000	05:18
Subject Sample # 2	0.162	05:19
Breath Volume	1.152	05:19
Air Blank	0.000	05:19
Control Test	0.077	05:20
Air Blank	0.000	05:20
Diagnostics Check	OK	05:20

Data Download Date/Time: 10/04/2018 11:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNYDER,P
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 05:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:14

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 09/23/2018 06:41:19

Diagnostics Check	OK	06:34
Air Blank	0.000	06:35
Control Test	0.079	06:35
Air Blank	0.000	06:36
Subject Sample # 1	0.112	06:36
Breath Volume	3.832	06:36
Air Blank	0.000	06:37
Air Blank	0.000	06:39
Subject Sample # 2	0.107	06:39
Breath Volume	4.042	06:39
Air Blank	0.000	06:40
Control Test	0.079	06:40
Air Blank	0.000	06:41
Diagnostics Check	OK	06:41

Data Download Date/Time: 10/04/2018 11:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: RAMOS YAMERA

DATE/TIME OF TEST: 09/25/2018 02:45:42

Diagnostics Check	OK	02:39
Air Blank	0.000	02:39
Control Test	0.080	02:39
Air Blank	0.000	02:40
Subject Sample # 1	0.093	02:41
Breath Volume	2.433	02:41
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.099	02:44
Breath Volume	1.886	02:44
Air Blank	0.000	02:44
Control Test	0.079	02:45
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 10/04/2018 11:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: STONESIFER

DATE/TIME OF TEST: 09/25/2018 04:18:16

Diagnostics Check	OK	04:11
Air Blank	0.000	04:12
Control Test	0.080	04:12
Air Blank	0.000	04:13
Subject Sample # 1	0.069	04:13
Breath Volume	2.101	04:13
Air Blank	0.000	04:14
Air Blank	0.000	04:16
Subject Sample # 2	0.071	04:16
Breath Volume	2.257	04:16
Air Blank	0.000	04:17
Control Test	0.079	04:17
Air Blank	0.000	04:18
Diagnostics Check	OK	04:18

Data Download Date/Time: 10/04/2018 11:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD, J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: BOYINGTON

DATE/TIME OF TEST: 09/26/2018 04:22:10

Diagnostics Check	OK	04:15
Air Blank	0.000	04:16
Control Test	0.079	04:16
Air Blank	0.000	04:16
Subject Sample # 1	0.000	04:17
Breath Volume	1.195	04:17
Air Blank	0.000	04:18
Air Blank	0.000	04:20
Subject Sample # 2	0.000	04:20
Breath Volume	1.468	04:20
Air Blank	0.000	04:21
Control Test	0.078	04:21
Air Blank	0.000	04:22
Diagnostics Check	OK	04:22

Data Download Date/Time: 10/04/2018 11:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ, J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 09/27/2018 02:32:29

Diagnostics Check	OK	02:21	
Air Blank	0.000	02:21	
Control Test	0.080	02:21	
Air Blank	0.000	02:22	
Subject Sample	# 1	02:25	No Sample Provided
Breath Volume	0.000	02:25	
Air Blank	0.000	02:25	
Air Blank	0.000	02:27	
Subject Sample	# 2	02:31	No Sample Provided
Breath Volume	0.000	02:31	
Air Blank	0.000	02:31	
Control Test	0.080	02:31	
Air Blank	0.000	02:32	
Diagnostics Check	OK	02:32	

Data Download Date/Time: 10/04/2018 11:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MILLER, J
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: ARCAROLA

DATE/TIME OF TEST: 09/27/2018 03:32:09

Diagnostics Check	OK	03:25
Air Blank	0.000	03:26
Control Test	0.079	03:26
Air Blank	0.000	03:27
Subject Sample # 1	0.000	03:27
Breath Volume	1.628	03:27
Air Blank	0.000	03:28
Air Blank	0.000	03:30
Subject Sample # 2	0.000	03:30
Breath Volume	1.421	03:30
Air Blank	0.000	03:31
Control Test	0.080	03:31
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

Data Download Date/Time: 10/04/2018 11:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PECYLAK,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:04

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 09/27/2018 05:36:37

Diagnostics Check	OK	05:29
Air Blank	0.000	05:30
Control Test	0.079	05:30
Air Blank	0.000	05:31
Subject Sample # 1	0.169	05:31
Breath Volume	2.808	05:31
Air Blank	0.000	05:32
Air Blank	0.000	05:34
Subject Sample # 2	0.173	05:35
Breath Volume	2.128	05:35
Air Blank	0.000	05:35
Control Test	0.079	05:36
Air Blank	0.000	05:36
Diagnostics Check	OK	05:36

Data Download Date/Time: 10/04/2018 11:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:27

SUBJECT'S LAST NAME: REGISTER III

DATE/TIME OF TEST: 09/27/2018 21:55:26

Diagnostics Check	OK	21:47
Air Blank	0.000	21:47
Control Test	0.080	21:48
Air Blank	0.000	21:48
Subject Sample # 1	0.084	21:51
Breath Volume	2.996	21:51
Air Blank	0.000	21:51
Air Blank	0.000	21:53
Subject Sample # 2	0.093	21:53
Breath Volume	2.894	21:53
Air Blank	0.000	21:54
Control Test	0.080	21:54
Air Blank	0.000	21:55
Diagnostics Check	OK	21:55

Data Download Date/Time: 10/04/2018 11:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANGELICCHI,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: BOUABID

DATE/TIME OF TEST: 09/29/2018 00:22:10

Diagnostics Check	OK	00:15
Air Blank	0.000	00:15
Control Test	0.080	00:16
Air Blank	0.000	00:16
Subject Sample # 1	0.197	00:17
Breath Volume	2.531	00:17
Air Blank	0.000	00:18
Air Blank	0.000	00:19
Subject Sample # 2	0.187	00:20
Breath Volume	1.722	00:20
Air Blank	0.000	00:21
Control Test	0.079	00:21
Air Blank	0.000	00:22
Diagnostics Check	OK	00:22

Data Download Date/Time: 10/04/2018 11:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FLORES,C
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 09/29/2018 04:58:49

Diagnostics Check	OK	04:51
Air Blank	0.000	04:52
Control Test	0.079	04:52
Air Blank	0.000	04:52
Subject Sample # 1	0.203	04:53
Breath Volume	2.335	04:53
Air Blank	0.000	04:54
Air Blank	0.000	04:56
Subject Sample # 2	0.194	04:57
Breath Volume	1.980	04:57
Air Blank	0.000	04:57
Control Test	0.078	04:58
Air Blank	0.000	04:58
Diagnostics Check	OK	04:58

Data Download Date/Time: 10/04/2018 11:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LYON, M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: DUGAN

DATE/TIME OF TEST: 09/30/2018 00:25:09

Diagnostics Check	OK	00:18
Air Blank	0.000	00:19
Control Test	0.080	00:19
Air Blank	0.000	00:20
Subject Sample # 1	0.156	00:20
Breath Volume	4.257	00:20
Air Blank	0.000	00:21
Air Blank	0.000	00:23
Subject Sample # 2	0.162	00:23
Breath Volume	4.496	00:23
Air Blank	0.000	00:24
Control Test	0.079	00:24
Air Blank	0.000	00:25
Diagnostics Check	OK	00:25

Data Download Date/Time: 10/04/2018 11:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: MARTELLI

DATE/TIME OF TEST: 09/30/2018 02:53:12

Diagnostics Check	OK	02:45
Air Blank	0.000	02:46
Control Test	0.080	02:46
Air Blank	0.000	02:47
Subject Sample # 1	0.163	02:48
Breath Volume	2.625	02:48
Air Blank	0.000	02:49
Air Blank	0.000	02:50
Subject Sample # 2	0.170	02:51
Breath Volume	3.089	02:51
Air Blank	0.000	02:52
Control Test	0.079	02:52
Air Blank	0.000	02:53
Diagnostics Check	OK	02:53

Data Download Date/Time: 10/04/2018 11:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: ROJAS-RAMIREZ

DATE/TIME OF TEST: 09/30/2018 21:36:59

Diagnostics Check	OK	21:30
Air Blank	0.000	21:30
Control Test	0.080	21:30
Air Blank	0.000	21:31
Subject Sample # 1	0.188	21:32
Breath Volume	1.839	21:32
Air Blank	0.000	21:32
Air Blank	0.000	21:34
Subject Sample # 2	0.179	21:35
Breath Volume	2.339	21:35
Air Blank	0.000	21:36
Control Test	0.079	21:36
Air Blank	0.000	21:36
Diagnostics Check	OK	21:36

Data Download Date/Time: 10/04/2018 11:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,JAMES,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: JONES,JAMES
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 08/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: DUMAS

DATE/TIME OF TEST: 09/06/2018 02:57:55

Diagnostics Check	OK	02:51
Air Blank	0.000	02:52
Control Test	0.079	02:52
Air Blank	0.000	02:52
Subject Sample # 1	0.103	02:53
Breath Volume	2.175	02:53
Air Blank	0.000	02:54
Air Blank	0.000	02:55
Subject Sample # 2	0.109	02:56
Breath Volume	3.203	02:56
Air Blank	0.000	02:57
Control Test	0.079	02:57
Air Blank	0.000	02:57
Diagnostics Check	OK	02:57

Data Download Date/Time: 09/10/2018 14:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: JONES,JAMES
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 08/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 09/08/2018 01:45:16

Diagnostics Check	OK	01:31	
Air Blank	0.000	01:31	
Control Test	0.080	01:31	
Air Blank	0.000	01:32	
Subject Sample	# 1 0.141	01:35	
Breath Volume	2.484	01:35	
Air Blank	0.000	01:35	
Air Blank	0.000	01:37	
Subject Sample	# 2 0.130	01:40	Volume Not Met
Breath Volume	0.148	01:40	
Air Blank	0.000	01:41	
Air Blank	0.000	01:43	
Subject Sample	# 3 0.134	01:43	
Breath Volume	1.980	01:43	
Air Blank	0.000	01:44	
Control Test	0.080	01:44	
Air Blank	0.000	01:45	
Diagnostics Check	OK	01:45	

Data Download Date/Time: 09/10/2018 14:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: WEBER,JEFFREY
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 09/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: SIMS

DATE/TIME OF TEST: 09/28/2018 04:16:28

Diagnostics Check	OK	04:09
Air Blank	0.000	04:10
Control Test	0.079	04:10
Air Blank	0.000	04:11
Subject Sample # 1	0.109	04:11
Breath Volume	3.679	04:11
Air Blank	0.000	04:12
Air Blank	0.000	04:14
Subject Sample # 2	0.109	04:15
Breath Volume	2.753	04:15
Air Blank	0.000	04:15
Control Test	0.080	04:15
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

Data Download Date/Time: 10/02/2018 17:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: LYNN HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYFIELD, MATTHEW, E
BREATH TEST OPERATOR AGENCY: Lynn Haven PD
ARRESTING OFFICER: MAYFIELD, MATTHEW
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 08/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 11:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:20

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 09/05/2018 11:51:18

Diagnostics Check	OK	11:44
Air Blank	0.000	11:45
Control Test	0.078	11:45
Air Blank	0.000	11:46
Subject Sample # 1	0.000	11:46
Breath Volume	3.578	11:46
Air Blank	0.000	11:47
Air Blank	0.000	11:49
Subject Sample # 2	0.000	11:49
Breath Volume	2.359	11:49
Air Blank	0.000	11:50
Control Test	0.079	11:50
Air Blank	0.000	11:51
Diagnostics Check	OK	11:51

Data Download Date/Time: 09/12/2018 03:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUEHLER,DEP,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KOSTYAL,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: METCALF

DATE/TIME OF TEST: 09/09/2018 00:52:33

Diagnostics Check	OK	00:45
Air Blank	0.000	00:46
Control Test	0.080	00:46
Air Blank	0.000	00:47
Subject Sample # 1	0.210	00:47
Breath Volume	3.914	00:47
Air Blank	0.000	00:48
Air Blank	0.000	00:50
Subject Sample # 2	0.204	00:51
Breath Volume	4.832	00:51
Air Blank	0.000	00:51
Control Test	0.079	00:51
Air Blank	0.000	00:52
Diagnostics Check	OK	00:52

Data Download Date/Time: 11/01/2018 11:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONG,CHRISTIAN,A
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BUEHLER,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: LONG

DATE/TIME OF TEST: 09/28/2018 22:24:06

Diagnostics Check	OK	22:17
Air Blank	0.000	22:17
Control Test	0.080	22:18
Air Blank	0.000	22:18
Subject Sample # 1	0.245	22:19
Breath Volume	3.039	22:19
Air Blank	0.000	22:20
Air Blank	0.000	22:22
Subject Sample # 2	0.237	22:22
Breath Volume	1.789	22:22
Air Blank	0.000	22:23
Control Test	0.079	22:23
Air Blank	0.000	22:24
Diagnostics Check	OK	22:24

Data Download Date/Time: 11/01/2018 11:10