Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLER,DAVID,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: HARVEY,R.
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A3
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: BAYLISS

DATE/TIME OF TEST: 11/01/2018 03:06:18

Diagnostics Check		OK	03:00
Air Blank		0.000	03:00
Control Test		0.079	03:00
Air Blank		0.000	03:01
Subject Sample	# 1	0.156	03:01
Breath Volume		2.191	03:01
Air Blank		0.000	03:02
Air Blank		0.000	03:04
Subject Sample	# 2	0.147	03:04
Breath Volume		1.546	03:04
Air Blank		0.000	03:05
Control Test		0.079	03:05
Air Blank		0.000	03:06
Diagnostics Check		OK	03:06

Data Download Date/Time: 11/13/2018 10:53

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAXTON,D,L
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: BRAXTON,D
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A3
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: BOSQUE

DATE/TIME OF TEST: 11/01/2018 03:15:39

Diagnostics Check		OK	03:09
Air Blank		0.000	03:09
Control Test		0.079	03:10
Air Blank		0.000	03:10
Subject Sample	# 1	0.136	03:11
Breath Volume		2.152	03:11
Air Blank		0.000	03:11
Air Blank		0.000	03:13
Subject Sample	# 2	0.140	03:14
Breath Volume		2.046	03:14
Air Blank		0.000	03:14
Control Test		0.078	03:15
Air Blank		0.000	03:15
Diagnostics Check		OK	03:15

Data Download Date/Time: 11/13/2018 10:53

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLER,DAVID,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: KELLER,DAVID
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A3
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: ANDRES-MARTINEZ

DATE/TIME OF TEST: 11/03/2018 04:16:18

Diagnostics Check		OK	04:10
Air Blank		0.000	04:10
Control Test		0.080	04:11
Air Blank		0.000	04:11
Subject Sample	# 1	0.240	04:12
Breath Volume		2.640	04:12
Air Blank		0.000	04:12
Air Blank		0.000	04:14
Subject Sample	# 2	0.246	04:14
Breath Volume		2.671	04:14
Air Blank		0.000	04:15
Control Test		0.079	04:15
Air Blank		0.000	04:16
Diagnostics Check		OK	04:16

Data Download Date/Time: 11/13/2018 10:53

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAN,DONALD,E
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: DEAN,DONALD
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A3
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: SLAYMAKER

DATE/TIME OF TEST: 11/03/2018 04:28:40

Diagnostics Check		OK	04:25	
Air Blank		0.000	04:26	
Control Test		0.079	04:26	
Air Blank		0.000	04:27	
Subject Sample	# 1		04:27	Subject Test Refused
Breath Volume		0.000	04:27	
Air Blank		0.000	04:27	
Control Test		0.080	04:28	
Air Blank		0.000	04:28	
Diagnostics Check		OK	04:28	

Data Download Date/Time: 11/13/2018 10:53

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLER,DAVID,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: KELLER,DAVID
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A3
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: KARDATZKE

DATE/TIME OF TEST: 11/04/2018 02:50:37

Diagnostics Check		OK	02:44
Air Blank		0.000	02:45
Control Test		0.080	02:45
Air Blank		0.000	02:45
Subject Sample	# 1	0.131	02:46
Breath Volume		1.859	02:46
Air Blank		0.000	02:46
Air Blank		0.000	02:48
Subject Sample	# 2	0.128	02:49
Breath Volume		1.523	02:49
Air Blank		0.000	02:49
Control Test		0.080	02:50
Air Blank		0.000	02:50
Diagnostics Check		OK	02:50

Data Download Date/Time: 11/13/2018 10:53

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLER,DAVID,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: COLLINS,C
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A3
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: CHAMBERLIN

DATE/TIME OF TEST: 11/04/2018 03:00:31

Diagnostics Check		OK	02:57	
Air Blank		0.000	02:58	
Control Test		0.080	02:58	
Air Blank		0.000	02:59	
Subject Sample	# 1		02:59	Subject Test Refused
Breath Volume		0.000	02:59	
Air Blank		0.000	02:59	
Control Test		0.080	02:59	
Air Blank		0.000	03:00	
Diagnostics Check		OK	03:00	

Data Download Date/Time: 11/13/2018 10:53

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAN,DONALD,E
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: DEAN,DONALD
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A3
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: GLAZE

DATE/TIME OF TEST: 11/04/2018 03:09:33

Diagnostics Check		OK	03:05	
Air Blank		0.000	03:06	
Control Test		0.078	03:06	
Air Blank		0.000	03:07	
Subject Sample	# 1		03:08	Sequence Aborted
Breath Volume		0.355	03:08	
Air Blank		0.000	03:09	

Data Download Date/Time: 11/13/2018 10:53

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BULLOCK,JOSHUA,R
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: COLLINS,T
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A3
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: SIMS

DATE/TIME OF TEST: 11/12/2018 02:51:05

Diagnostics Check		OK	02:44
Air Blank		0.000	02:45
Control Test		0.079	02:45
Air Blank		0.000	02:46
Subject Sample	# 1	0.194	02:46
Breath Volume		1.726	02:46
Air Blank		0.000	02:47
Air Blank		0.000	02:49
Subject Sample	# 2	0.191	02:49
Breath Volume		1.828	02:49
Air Blank		0.000	02:50
Control Test		0.080	02:50
Air Blank		0.000	02:50
Diagnostics Check		OK	02:51

Data Download Date/Time: 11/13/2018 10:53

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIXON,WESLEY,B
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: DIXON,WESLEY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 11/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A3
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:25

SUBJECT'S LAST NAME: CLINE

DATE/TIME OF TEST: 11/13/2018 17:55:33

Diagnostics Check		OK	17:49
Air Blank		0.000	17:49
Control Test		0.079	17:49
Air Blank		0.000	17:50
Subject Sample	# 1	0.202	17:50
Breath Volume		3.464	17:50
Air Blank		0.000	17:51
Air Blank		0.000	17:53
Subject Sample	# 2	0.199	17:54
Breath Volume		3.523	17:54
Air Blank		0.000	17:54
Control Test		0.078	17:54
Air Blank		0.000	17:55
Diagnostics Check		OK	17:55

Data Download Date/Time: 12/14/2018 10:03

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLER,DAVID,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: KELLER,DAVID
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 11/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A3
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 11/15/2018 00:03:07

Diagnostics Check		OK	23:56
Air Blank		0.000	23:56
Control Test		0.080	23:57
Air Blank		0.000	23:57
Subject Sample	# 2	0.064	23:58
Breath Volume		1.417	23:58
Air Blank		0.000	23:58
Air Blank		0.000	00:00
Subject Sample	# 1	0.063	00:01
Breath Volume		1.406	00:01
Air Blank		0.000	00:02
Control Test		0.080	00:02
Air Blank		0.000	00:03
Diagnostics Check		OK	00:03

Data Download Date/Time: 12/14/2018 10:03

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEELE,R,L
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: STEELE,R
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 11/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A3
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: GIBSON

DATE/TIME OF TEST: 11/15/2018 03:57:34

Diagnostics Check		OK	03:50
Air Blank		0.000	03:50
Control Test		0.079	03:50
Air Blank		0.000	03:51
Subject Sample	# 1	0.149	03:52
Breath Volume		1.925	03:52
Air Blank		0.000	03:52
Air Blank		0.000	03:54
Subject Sample	# 2	0.145	03:56
Breath Volume		2.007	03:56
Air Blank		0.000	03:56
Control Test		0.079	03:56
Air Blank		0.000	03:57
Diagnostics Check		OK	03:57

Data Download Date/Time: 12/14/2018 10:03

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAXTON,D,L
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ARMSTRONG,J
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 11/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A3
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: CRUMEL

DATE/TIME OF TEST: 11/16/2018 03:06:16

Diagnostics Check		OK	02:56
Air Blank		0.000	02:57
Control Test		0.080	02:57
Air Blank		0.000	02:58
Subject Sample	# 1	0.163	02:59
Breath Volume		1.695	02:59
Air Blank		0.000	03:00
Air Blank		0.000	03:01
Subject Sample	# 2	0.160	03:04
Breath Volume		1.851	03:04
Air Blank		0.000	03:05
Control Test		0.079	03:05
Air Blank		0.000	03:06
Diagnostics Check		OK	03:06

Data Download Date/Time: 12/14/2018 10:03

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,JEFFREY,A
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: HARVEY,RHETT
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 11/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A3
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: RICKETTS

DATE/TIME OF TEST: 11/17/2018 02:19:43

Diagnostics Check		OK	02:06	
Air Blank		0.000	02:06	
Control Test		0.081	02:07	
Air Blank		0.000	02:07	
Subject Sample	# 1	0.194	02:09	No .020 agreement
Breath Volume		4.757	02:09	
Air Blank		0.000	02:09	
Air Blank		0.000	02:11	
Subject Sample	# 2		02:14	No Sample Provided
Breath Volume		0.000	02:14	
Air Blank		0.000	02:15	
Air Blank		0.000	02:17	
Subject Sample	# 3	0.164	02:18	No .020 agreement
Breath Volume		1.703	02:18	
Air Blank		0.000	02:18	
Control Test		0.079	02:19	
Air Blank		0.000	02:19	
Diagnostics Check		OK	02:19	

Data Download Date/Time: 12/14/2018 10:03

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAS,BRANDYN,M
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: DEAS,BRANDYN
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 11/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A3
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 13:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:40

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 11/17/2018 15:07:55

Diagnostics Check		OK	15:01
Air Blank		0.000	15:01
Control Test		0.081	15:01
Air Blank		0.000	15:02
Subject Sample	# 1	0.011	15:02
Breath Volume		1.785	15:02
Air Blank		0.000	15:03
Air Blank		0.000	15:05
Subject Sample	# 2	0.010	15:06
Breath Volume		1.734	15:06
Air Blank		0.000	15:07
Control Test		0.081	15:07
Air Blank		0.000	15:07
Diagnostics Check		OK	15:07

Data Download Date/Time: 12/14/2018 10:03

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEELE,R,L
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: COLLINS,C
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 11/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A3
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: COFER

DATE/TIME OF TEST: 11/21/2018 03:01:48

Diagnostics Check		OK	02:55
Air Blank		0.000	02:55
Control Test		0.080	02:56
Air Blank		0.000	02:56
Subject Sample	# 1	0.292	02:57
Breath Volume		2.238	02:57
Air Blank		0.000	02:58
Air Blank		0.000	02:59
Subject Sample	# 2	0.280	03:00
Breath Volume		2.144	03:00
Air Blank		0.000	03:00
Control Test		0.078	03:01
Air Blank		0.000	03:01
Diagnostics Check		OK	03:01

Data Download Date/Time: 12/14/2018 10:03

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAN,DONALD,E
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: DEAN,DONALD
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 11/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A3
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: BLANKINSHIP

DATE/TIME OF TEST: 11/21/2018 03:33:37

Diagnostics Check	OK	03:33	
Air Blank		03:33	RFI Detect
Air Blank	0.000	03:33	

Data Download Date/Time: 12/14/2018 10:03

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PILKANIS,MARY,P
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: ELDRIDGE,JARROD
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 11/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A3
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 11/22/2018 00:19:52

Diagnostics Check		OK	00:12
Air Blank		0.000	00:12
Control Test		0.080	00:13
Air Blank		0.000	00:13
Subject Sample	# 1	0.224	00:15
Breath Volume		1.832	00:15
Air Blank		0.000	00:15
Air Blank		0.000	00:17
Subject Sample	# 2	0.218	00:18
Breath Volume		1.433	00:18
Air Blank		0.000	00:18
Control Test		0.079	00:19
Air Blank		0.000	00:19
Diagnostics Check		OK	00:19

Data Download Date/Time: 12/14/2018 10:03

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,G,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,G
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 11/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A3
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: BALRAJ

DATE/TIME OF TEST: 11/22/2018 01:50:29

Diagnostics Check		OK	01:44
Air Blank		0.000	01:44
Control Test		0.080	01:45
Air Blank		0.000	01:45
Subject Sample	# 1	0.000	01:46
Breath Volume		3.144	01:46
Air Blank		0.000	01:46
Air Blank		0.000	01:48
Subject Sample	# 2	0.000	01:49
Breath Volume		1.300	01:49
Air Blank		0.000	01:49
Control Test		0.081	01:49
Air Blank		0.000	01:50
Diagnostics Check		OK	01:50

Data Download Date/Time: 12/14/2018 10:03

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHIPMAN,KAREN,D
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: SHIPMAN,KAREN
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 11/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A3
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 14:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:19

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 11/23/2018 15:56:48

Diagnostics Check		OK	15:50
Air Blank		0.000	15:51
Control Test		0.079	15:51
Air Blank		0.000	15:52
Subject Sample	# 1	0.206	15:52
Breath Volume		1.324	15:52
Air Blank		0.000	15:53
Air Blank		0.000	15:54
Subject Sample	# 2	0.203	15:55
Breath Volume		1.437	15:55
Air Blank		0.000	15:55
Control Test		0.079	15:56
Air Blank		0.000	15:56
Diagnostics Check		OK	15:56

Data Download Date/Time: 12/14/2018 10:03

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAXTON,D,L
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: BLEVINS,A
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 11/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A3
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: TWYMAN

DATE/TIME OF TEST: 11/25/2018 03:35:59

Diagnostics Check		OK	03:29
Air Blank		0.000	03:30
Control Test		0.079	03:30
Air Blank		0.000	03:31
Subject Sample	# 1	0.143	03:31
Breath Volume		3.156	03:31
Air Blank		0.000	03:32
Air Blank		0.000	03:34
Subject Sample	# 2	0.143	03:34
Breath Volume		2.281	03:34
Air Blank		0.000	03:35
Control Test		0.079	03:35
Air Blank		0.000	03:35
Diagnostics Check		OK	03:35

Data Download Date/Time: 12/14/2018 10:03

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARNER,TYLER,J
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: WEST,JAMES
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 11/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A3
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: NEAL

DATE/TIME OF TEST: 11/29/2018 23:22:55

Diagnostics Check		OK	23:16
Air Blank		0.000	23:16
Control Test		0.081	23:16
Air Blank		0.000	23:17
Subject Sample	# 1	0.000	23:18
Breath Volume		1.601	23:18
Air Blank		0.000	23:19
Air Blank		0.000	23:21
Subject Sample	# 2	0.000	23:21
Breath Volume		1.382	23:21
Air Blank		0.000	23:22
Control Test		0.081	23:22
Air Blank		0.000	23:22
Diagnostics Check		OK	23:22

Data Download Date/Time: 12/14/2018 10:03

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTHWAY,GEORGIA,K
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: NORTHWAY,GEORGIA
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 11/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A3
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 11/30/2018 00:06:05

Diagnostics Check		OK	00:01	
Air Blank		0.000	00:01	
Control Test		0.080	00:01	
Air Blank		0.000	00:02	
Subject Sample	# 1		00:04	Subject Test Refused
Breath Volume		0.859	00:04	
Air Blank		0.000	00:05	
Control Test		0.080	00:05	
Air Blank		0.000	00:06	
Diagnostics Check		OK	00:06	

Data Download Date/Time: 12/14/2018 10:03

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLER,DAVID,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: STEELE,R
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 11/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A3
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: DAY

DATE/TIME OF TEST: 11/30/2018 03:21:47

Diagnostics Check		OK	03:15
Air Blank		0.000	03:16
Control Test		0.080	03:16
Air Blank		0.000	03:16
Subject Sample	# 1	0.197	03:17
Breath Volume		1.582	03:17
Air Blank		0.000	03:18
Air Blank		0.000	03:19
Subject Sample	# 2	0.203	03:20
Breath Volume		1.289	03:20
Air Blank		0.000	03:20
Control Test		0.080	03:21
Air Blank		0.000	03:21
Diagnostics Check		OK	03:21

Data Download Date/Time: 12/14/2018 10:03

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLER,DAVID,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: COLLINS,C
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 11/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A3
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: ESPINOSA

DATE/TIME OF TEST: 11/30/2018 03:36:27

Diagnostics Check		OK	03:28
Air Blank		0.000	03:28
Control Test		0.080	03:29
Air Blank		0.000	03:29
Subject Sample	# 1	0.183	03:31
Breath Volume		1.753	03:31
Air Blank		0.000	03:32
Air Blank		0.000	03:34
Subject Sample	# 2	0.186	03:35
Breath Volume		1.351	03:35
Air Blank		0.000	03:35
Control Test		0.079	03:35
Air Blank		0.000	03:36
Diagnostics Check		OK	03:36

Data Download Date/Time: 12/14/2018 10:03

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEIERS,BLAINE,T
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: WEIERS,BLAINE
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 10/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: MACLEOD

DATE/TIME OF TEST: 11/04/2018 01:44:06

Diagnostics Check		OK	01:37
Air Blank		0.000	01:37
Control Test		0.080	01:37
Air Blank		0.000	01:38
Subject Sample	# 1	0.232	01:39
Breath Volume		1.484	01:39
Air Blank		0.000	01:40
Air Blank		0.000	01:41
Subject Sample	# 2	0.238	01:42
Breath Volume		1.593	01:42
Air Blank		0.000	01:43
Control Test		0.079	01:43
Air Blank		0.000	01:44
Diagnostics Check		OK	01:44

Data Download Date/Time: 11/12/2018 23:20

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLER,DAVID,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: KELLER,DAVID
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: DUNBAR

DATE/TIME OF TEST: 11/02/2018 03:50:33

Diagnostics Check		OK	03:43
Air Blank		0.000	03:43
Control Test		0.079	03:44
Air Blank		0.000	03:44
Subject Sample	# 1	0.257	03:46
Breath Volume		1.503	03:46
Air Blank		0.000	03:46
Air Blank		0.000	03:48
Subject Sample	# 2	0.254	03:49
Breath Volume		1.683	03:49
Air Blank		0.000	03:49
Control Test		0.077	03:49
Air Blank		0.000	03:50
Diagnostics Check		OK	03:50

Data Download Date/Time: 11/13/2018 10:56

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAN,DONALD,E
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: DEAN,DONALD
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: GLAZE

DATE/TIME OF TEST: 11/04/2018 03:23:10

Diagnostics Check		OK	03:16
Air Blank		0.000	03:17
Control Test		0.079	03:17
Air Blank		0.000	03:18
Subject Sample	# 1	0.187	03:18
Breath Volume		1.867	03:18
Air Blank		0.000	03:19
Air Blank		0.000	03:21
Subject Sample	# 2	0.193	03:21
Breath Volume		3.218	03:21
Air Blank		0.000	03:22
Control Test		0.078	03:22
Air Blank		0.000	03:23
Diagnostics Check		OK	03:23

Data Download Date/Time: 11/13/2018 10:56

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLER,DAVID,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: KELLER,DAVID
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: LOWE

DATE/TIME OF TEST: 11/08/2018 03:34:24

Diagnostics Check		OK	03:28
Air Blank		0.000	03:28
Control Test		0.080	03:29
Air Blank		0.000	03:29
Subject Sample	# 1	0.193	03:30
Breath Volume		2.644	03:30
Air Blank		0.000	03:30
Air Blank		0.000	03:32
Subject Sample	# 2	0.201	03:32
Breath Volume		2.546	03:32
Air Blank		0.000	03:33
Control Test		0.079	03:33
Air Blank		0.000	03:34
Diagnostics Check		OK	03:34

Data Download Date/Time: 11/13/2018 10:56

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIXON,WESLEY,B
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: DIXON,WESLEY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 11/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:25

SUBJECT'S LAST NAME: CLINE

DATE/TIME OF TEST: 11/13/2018 18:00:58

Diagnostics Check	OK	17:59	
Air Blank	0.000	17:59	
Control Test	0.079	18:00	
Air Blank		18:00	RFI Detect
Air Blank	0.000	18:00	

Data Download Date/Time: 12/14/2018 10:06

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIXON,WESLEY,B
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: DIXON,WESLEY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 11/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:25

SUBJECT'S LAST NAME: CLINE

DATE/TIME OF TEST: 11/13/2018 18:09:17

Diagnostics Check		OK	18:03
Air Blank		0.000	18:03
Control Test		0.079	18:03
Air Blank		0.000	18:04
Subject Sample	# 1	0.198	18:04
Breath Volume		3.324	18:04
Air Blank		0.000	18:05
Air Blank		0.000	18:07
Subject Sample	# 2	0.190	18:07
Breath Volume		3.265	18:07
Air Blank		0.000	18:08
Control Test		0.078	18:08
Air Blank		0.000	18:09
Diagnostics Check		OK	18:09

Data Download Date/Time: 12/14/2018 10:06

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAN,DONALD,E
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: DEAN,DONALD
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 11/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: BLANKINSHIP

DATE/TIME OF TEST: 11/21/2018 03:36:45

Diagnostics Check		OK	03:33	
Air Blank		0.000	03:34	
Control Test		0.081	03:34	
Air Blank		0.000	03:35	
Subject Sample	# 1		03:35	Subject Test Refused
Breath Volume		0.000	03:35	
Air Blank		0.000	03:35	
Control Test		0.081	03:36	
Air Blank		0.000	03:36	
Diagnostics Check		OK	03:36	

Data Download Date/Time: 12/14/2018 10:06

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,G,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,G
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 11/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: SWERDLEN

DATE/TIME OF TEST: 11/24/2018 00:03:40

Diagnostics Check		OK	23:57
Air Blank		0.000	23:57
Control Test		0.081	23:58
Air Blank		0.000	23:58
Subject Sample	# 2	0.058	23:59
Breath Volume		1.726	23:59
Air Blank		0.000	23:59
Air Blank		0.000	00:01
Subject Sample	# 1	0.061	00:02
Breath Volume		2.382	00:02
Air Blank		0.000	00:02
Control Test		0.081	00:03
Air Blank		0.000	00:03
Diagnostics Check		OK	00:03

Data Download Date/Time: 12/14/2018 10:06

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000228

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASTINGS,KENNETH,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: HASTINGS,KENNETH
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: VINDAS

DATE/TIME OF TEST: 11/01/2018 04:48:01

Diagnostics Check		OK	04:41
Air Blank		0.000	04:41
Control Test		0.081	04:41
Air Blank		0.000	04:42
Subject Sample	# 1	0.135	04:43
Breath Volume		1.613	04:43
Air Blank		0.000	04:43
Air Blank		0.000	04:45
Subject Sample	# 2	0.142	04:46
Breath Volume		1.855	04:46
Air Blank		0.000	04:47
Control Test		0.080	04:47
Air Blank		0.000	04:47
Diagnostics Check		OK	04:48

Data Download Date/Time: 11/01/2018 14:11

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000228

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: FERREYRA

DATE/TIME OF TEST: 11/01/2018 21:00:23

Diagnostics Check		OK	20:52
Air Blank		0.000	20:52
Control Test		0.080	20:52
Air Blank		0.000	20:53
Subject Sample	# 1	0.146	20:55
Breath Volume		1.527	20:55
Air Blank		0.000	20:56
Air Blank		0.000	20:58
Subject Sample	# 2	0.147	20:58
Breath Volume		1.757	20:58
Air Blank		0.000	20:59
Control Test		0.079	20:59
Air Blank		0.000	21:00
Diagnostics Check		OK	21:00

Data Download Date/Time: 12/20/2018 12:58

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000228

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASTINGS,KENNETH,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: HASTINGS,KENNETH
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: JANKOWSKI

DATE/TIME OF TEST: 11/03/2018 02:54:15

Diagnostics Check		OK	02:50	
Air Blank		0.000	02:50	
Control Test		0.081	02:51	
Air Blank		0.000	02:51	
Subject Sample	# 1		02:52	Subject Test Refused
Breath Volume		0.000	02:52	
Air Blank		0.000	02:53	
Control Test		0.081	02:53	
Air Blank		0.000	02:54	
Diagnostics Check		OK	02:54	

Data Download Date/Time: 12/20/2018 12:58

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000228

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALAMEDA,JORGE,L
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: ALAMEDA,JORGE
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:40

SUBJECT'S LAST NAME: REESE

DATE/TIME OF TEST: 11/05/2018 17:12:05

Diagnostics Check		OK	17:05
Air Blank		0.000	17:06
Control Test		0.080	17:06
Air Blank		0.000	17:07
Subject Sample	# 1	0.207	17:07
Breath Volume		3.207	17:07
Air Blank		0.000	17:08
Air Blank		0.000	17:09
Subject Sample	# 2	0.206	17:10
Breath Volume		3.070	17:10
Air Blank		0.000	17:11
Control Test		0.079	17:11
Air Blank		0.000	17:12
Diagnostics Check		OK	17:12

Data Download Date/Time: 12/20/2018 12:58

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000228

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BALDIE,BRYAN,A
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: BALDIE,BRYAN
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 11/06/2018 19:00:29

Diagnostics Check		OK	18:52	
Air Blank		0.000	18:52	
Control Test		0.080	18:52	
Air Blank		0.000	18:53	
Subject Sample	# 1		18:56	No Sample Provided
Breath Volume		0.000	18:56	
Air Blank		0.000	18:56	
Air Blank		0.000	18:58	
Subject Sample	# 2		18:59	Subject Test Refused
Breath Volume		0.000	18:59	
Air Blank		0.000	18:59	
Control Test		0.080	18:59	
Air Blank		0.000	19:00	
Diagnostics Check		OK	19:00	

Data Download Date/Time: 12/20/2018 12:58

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000228

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASTINGS,KENNETH,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: HASTINGS,KENNETH
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 11/10/2018 22:34:36

Diagnostics Check		OK	22:28
Air Blank		0.000	22:28
Control Test		0.080	22:29
Air Blank		0.000	22:29
Subject Sample	# 1	0.136	22:30
Breath Volume		2.902	22:30
Air Blank		0.000	22:30
Air Blank		0.000	22:32
Subject Sample	# 2	0.132	22:32
Breath Volume		2.875	22:32
Air Blank		0.000	22:33
Control Test		0.080	22:33
Air Blank		0.000	22:34
Diagnostics Check		OK	22:34

Data Download Date/Time: 12/20/2018 12:58

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000228

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNSBROUGH,DREW,A
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: THORNSBROUGH,DREW
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 10:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:45

SUBJECT'S LAST NAME: SCHIEL

DATE/TIME OF TEST: 11/12/2018 11:13:08

Diagnostics Check		OK	11:06
Air Blank		0.000	11:07
Control Test		0.080	11:07
Air Blank		0.000	11:08
Subject Sample	# 1	0.000	11:08
Breath Volume		1.398	11:08
Air Blank		0.000	11:09
Air Blank		0.000	11:11
Subject Sample	# 2	0.000	11:11
Breath Volume		1.730	11:11
Air Blank		0.000	11:12
Control Test		0.080	11:12
Air Blank		0.000	11:13
Diagnostics Check		OK	11:13

Data Download Date/Time: 12/20/2018 12:58

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000228

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,ROBERT,C
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: COOK,ROBERT
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 15:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:19

SUBJECT'S LAST NAME: MORIAN

DATE/TIME OF TEST: 11/15/2018 16:46:00

Diagnostics Check		OK	16:40
Air Blank		0.000	16:40
Control Test		0.080	16:40
Air Blank		0.000	16:41
Subject Sample	# 1	0.000	16:41
Breath Volume		1.839	16:41
Air Blank		0.000	16:42
Air Blank		0.000	16:44
Subject Sample	# 2	0.000	16:44
Breath Volume		2.242	16:44
Air Blank		0.000	16:45
Control Test		0.081	16:45
Air Blank		0.000	16:45
Diagnostics Check		OK	16:46

Data Download Date/Time: 12/20/2018 12:58

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000228

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASTINGS,KENNETH,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: HASTINGS,KENNETH
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 11/16/2018 02:59:59

Diagnostics Check	OK	02:59	
Air Blank		02:59	RFI Detect
Air Blank	0.000	02:59	

Data Download Date/Time: 12/20/2018 12:58

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000228

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASTINGS,KENNETH,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: HASTINGS,KENNETH
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 11/16/2018 03:07:53

Diagnostics Check		OK	03:01
Air Blank		0.000	03:02
Control Test		0.080	03:02
Air Blank		0.000	03:02
Subject Sample	# 1	0.139	03:03
Breath Volume		3.160	03:03
Air Blank		0.000	03:04
Air Blank		0.000	03:05
Subject Sample	# 2	0.139	03:06
Breath Volume		2.371	03:06
Air Blank		0.000	03:06
Control Test		0.079	03:07
Air Blank		0.000	03:07
Diagnostics Check		OK	03:07

Data Download Date/Time: 12/20/2018 12:58

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000228

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 11/17/2018 02:11:13

Diagnostics Check		OK	02:08	
Air Blank		0.000	02:08	
Control Test		0.081	02:09	
Air Blank		0.000	02:09	
Subject Sample	# 1		02:09	Subject Test Refused
Breath Volume		0.000	02:09	
Air Blank		0.000	02:10	
Control Test		0.081	02:10	
Air Blank		0.000	02:11	
Diagnostics Check		OK	02:11	

Data Download Date/Time: 12/20/2018 12:58

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000228

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: BRITTON

DATE/TIME OF TEST: 11/17/2018 22:07:47

Diagnostics Check		OK	22:00
Air Blank		0.000	22:01
Control Test		0.080	22:01
Air Blank		0.000	22:01
Subject Sample	# 1	0.000	22:03
Breath Volume		1.757	22:03
Air Blank		0.000	22:03
Air Blank		0.000	22:05
Subject Sample	# 2	0.000	22:06
Breath Volume		1.539	22:06
Air Blank		0.000	22:06
Control Test		0.080	22:07
Air Blank		0.000	22:07
Diagnostics Check		OK	22:07

Data Download Date/Time: 12/20/2018 12:58

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000228

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASTINGS,KENNETH,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: HASTINGS,KENNETH
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: CHAMBERLAIN

DATE/TIME OF TEST: 11/18/2018 00:11:07

Diagnostics Check		OK	00:07	
Air Blank		0.000	00:08	
Control Test		0.080	00:08	
Air Blank		0.000	00:09	
Subject Sample	# 1		00:09	Subject Test Refused
Breath Volume		0.000	00:09	
Air Blank		0.000	00:10	
Control Test		0.081	00:10	
Air Blank		0.000	00:11	
Diagnostics Check		OK	00:11	

Data Download Date/Time: 12/20/2018 12:58

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000228

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 11/18/2018 01:23:56

Diagnostics Check		OK	01:17
Air Blank		0.000	01:18
Control Test		0.080	01:18
Air Blank		0.000	01:18
Subject Sample	# 1	0.000	01:19
Breath Volume		2.617	01:19
Air Blank		0.000	01:19
Air Blank		0.000	01:21
Subject Sample	# 2	0.000	01:22
Breath Volume		2.292	01:22
Air Blank		0.000	01:23
Control Test		0.081	01:23
Air Blank		0.000	01:23
Diagnostics Check		OK	01:23

Data Download Date/Time: 12/20/2018 12:58

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000228

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASTINGS,KENNETH,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: HASTINGS,KENNETH
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: SIRMANS

DATE/TIME OF TEST: 11/18/2018 03:35:23

Diagnostics Check		OK	03:32	
Air Blank		0.000	03:33	
Control Test		0.081	03:33	
Air Blank		0.000	03:33	
Subject Sample	# 1		03:34	Subject Test Refused
Breath Volume		0.000	03:34	
Air Blank		0.000	03:34	
Control Test		0.080	03:34	
Air Blank		0.000	03:35	
Diagnostics Check		OK	03:35	

Data Download Date/Time: 12/20/2018 12:58

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000228

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASTINGS,KENNETH,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: HASTINGS,KENNETH
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: EDNEY

DATE/TIME OF TEST: 11/21/2018 23:25:32

Diagnostics Check		OK	23:19
Air Blank		0.000	23:19
Control Test		0.081	23:19
Air Blank		0.000	23:20
Subject Sample	# 1	0.131	23:20
Breath Volume		2.570	23:20
Air Blank		0.000	23:21
Air Blank		0.000	23:23
Subject Sample	# 2	0.128	23:23
Breath Volume		1.789	23:23
Air Blank		0.000	23:24
Control Test		0.080	23:24
Air Blank		0.000	23:25
Diagnostics Check		OK	23:25

Data Download Date/Time: 12/20/2018 12:59

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000228

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: PRIDE

DATE/TIME OF TEST: 11/22/2018 00:50:44

Diagnostics Check		OK	00:43
Air Blank		0.000	00:44
Control Test		0.080	00:44
Air Blank		0.000	00:44
Subject Sample	# 1	0.063	00:46
Breath Volume		1.992	00:46
Air Blank		0.000	00:46
Air Blank		0.000	00:48
Subject Sample	# 2	0.062	00:49
Breath Volume		1.609	00:49
Air Blank		0.000	00:49
Control Test		0.080	00:50
Air Blank		0.000	00:50
Diagnostics Check		OK	00:50

Data Download Date/Time: 12/20/2018 12:59

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000228

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: BECHTOLD

DATE/TIME OF TEST: 11/23/2018 21:30:41

Diagnostics Check		OK	21:18	
Air Blank		0.000	21:18	
Control Test		0.080	21:19	
Air Blank		0.000	21:19	
Subject Sample	# 1		21:22	No Sample Provided
Breath Volume		0.000	21:22	
Air Blank		0.000	21:23	
Air Blank		0.000	21:25	
Subject Sample	# 2	0.139	21:25	
Breath Volume		2.863	21:25	
Air Blank		0.000	21:26	
Air Blank		0.000	21:28	
Subject Sample	# 3	0.135	21:29	
Breath Volume		2.832	21:29	
Air Blank		0.000	21:29	
Control Test		0.079	21:30	
Air Blank		0.000	21:30	
Diagnostics Check		OK	21:30	

Data Download Date/Time: 12/20/2018 12:59

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000228

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASTINGS,KENNETH,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: HASTINGS,KENNETH
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: BROGHAMER

DATE/TIME OF TEST: 11/24/2018 00:54:24

Diagnostics Check		OK	00:51	
Air Blank		0.000	00:51	
Control Test		0.079	00:52	
Air Blank		0.000	00:52	
Subject Sample	# 1		00:53	Subject Test Refused
Breath Volume		0.000	00:53	
Air Blank		0.000	00:53	
Control Test		0.080	00:53	
Air Blank		0.000	00:54	
Diagnostics Check		OK	00:54	

Data Download Date/Time: 12/20/2018 12:59

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUNT,CHARLES,R
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: HUNT,CHARLES
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 10/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: OSBORNE

DATE/TIME OF TEST: 11/09/2018 02:22:01

Diagnostics Check		OK	02:15
Air Blank		0.000	02:15
Control Test		0.078	02:16
Air Blank		0.000	02:16
Subject Sample	# 1	0.085	02:17
Breath Volume		1.882	02:17
Air Blank		0.000	02:17
Air Blank		0.000	02:19
Subject Sample	# 2	0.089	02:20
Breath Volume		1.984	02:20
Air Blank		0.000	02:21
Control Test		0.079	02:21
Air Blank		0.000	02:21
Diagnostics Check		OK	02:22

Data Download Date/Time: 11/27/2018 10:10

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESNEUL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: ESNEUL,MICHAEL
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 10/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: HERRERA PENATE

DATE/TIME OF TEST: 11/03/2018 01:16:00

Diagnostics Check		OK	01:09
Air Blank		0.000	01:09
Control Test		0.080	01:09
Air Blank		0.000	01:10
Subject Sample	# 1	0.000	01:11
Breath Volume		1.152	01:11
Air Blank		0.000	01:11
Air Blank		0.000	01:13
Subject Sample	# 2	0.000	01:14
Breath Volume		2.007	01:14
Air Blank		0.000	01:15
Control Test		0.080	01:15
Air Blank		0.000	01:15
Diagnostics Check		OK	01:16

Data Download Date/Time: 12/04/2018 14:58

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOWNING,CADEN,T
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: DOWNING,CADEN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 10/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: ST CYR

DATE/TIME OF TEST: 11/03/2018 02:31:14

Diagnostics Check		OK	02:24
Air Blank		0.000	02:25
Control Test		0.080	02:25
Air Blank		0.000	02:26
Subject Sample	# 1	0.155	02:26
Breath Volume		1.621	02:26
Air Blank		0.000	02:27
Air Blank		0.000	02:29
Subject Sample	# 2	0.156	02:29
Breath Volume		1.804	02:29
Air Blank		0.000	02:30
Control Test		0.080	02:30
Air Blank		0.000	02:31
Diagnostics Check		OK	02:31

Data Download Date/Time: 12/04/2018 14:58

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DONOVAN,CHARLES,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SANDERSON,JOHN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 10/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 06:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:30

SUBJECT'S LAST NAME: MRUZ

DATE/TIME OF TEST: 11/03/2018 06:58:35

Diagnostics Check		OK	06:51
Air Blank		0.000	06:51
Control Test		0.079	06:52
Air Blank		0.000	06:52
Subject Sample	# 1	0.188	06:53
Breath Volume		2.156	06:53
Air Blank		0.000	06:54
Air Blank		0.000	06:56
Subject Sample	# 2	0.171	06:56
Breath Volume		3.117	06:56
Air Blank		0.000	06:57
Control Test		0.079	06:57
Air Blank		0.000	06:58
Diagnostics Check		OK	06:58

Data Download Date/Time: 12/04/2018 14:58

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KANE,MARK,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: KANE,MARK
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 10/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 04:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: TALAVERA SANTIAGO

DATE/TIME OF TEST: 11/10/2018 05:23:13

Diagnostics Check		OK	05:15
Air Blank		0.000	05:16
Control Test		0.080	05:16
Air Blank		0.000	05:17
Subject Sample	# 1	0.163	05:18
Breath Volume		3.347	05:18
Air Blank		0.000	05:18
Air Blank		0.000	05:20
Subject Sample	# 2	0.149	05:21
Breath Volume		1.449	05:21
Air Blank		0.000	05:22
Control Test		0.079	05:22
Air Blank		0.000	05:23
Diagnostics Check		OK	05:23

Data Download Date/Time: 12/04/2018 14:58

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIMBRIL,JOSHUA,B
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: KIMBRIL,JOSHUA
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 10/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 17:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: FREEMAN

DATE/TIME OF TEST: 11/10/2018 18:17:15

Diagnostics Check		OK	18:10
Air Blank		0.000	18:10
Control Test		0.080	18:11
Air Blank		0.000	18:11
Subject Sample	# 1	0.150	18:12
Breath Volume		1.933	18:12
Air Blank		0.000	18:13
Air Blank		0.000	18:15
Subject Sample	# 2	0.146	18:15
Breath Volume		1.191	18:15
Air Blank		0.000	18:16
Control Test		0.080	18:16
Air Blank		0.000	18:17
Diagnostics Check		OK	18:17

Data Download Date/Time: 12/04/2018 14:58

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLE,JACKIE,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LOPEZ,CARLOS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 11:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:05

SUBJECT'S LAST NAME: COLE

DATE/TIME OF TEST: 11/13/2018 12:38:25

Diagnostics Check		OK	12:32
Air Blank		0.000	12:32
Control Test		0.079	12:32
Air Blank		0.000	12:33
Subject Sample	# 1	0.075	12:34
Breath Volume		2.320	12:34
Air Blank		0.000	12:34
Air Blank		0.000	12:36
Subject Sample	# 2	0.072	12:36
Breath Volume		2.558	12:36
Air Blank		0.000	12:37
Control Test		0.079	12:37
Air Blank		0.000	12:38
Diagnostics Check		OK	12:38

Data Download Date/Time: 12/04/2018 14:58

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLAUSNER,ALEXANDRA,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: CLAUSNER,ALEXANDRA
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 10/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: SYKES

DATE/TIME OF TEST: 11/16/2018 23:59:22

Diagnostics Check		OK	23:53
Air Blank		0.000	23:53
Control Test		0.080	23:53
Air Blank		0.000	23:54
Subject Sample	# 1	0.027	23:54
Breath Volume		1.796	23:54
Air Blank		0.000	23:55
Air Blank		0.000	23:57
Subject Sample	# 2	0.019	23:57
Breath Volume		2.226	23:57
Air Blank		0.000	23:58
Control Test		0.080	23:58
Air Blank		0.000	23:59
Diagnostics Check		OK	23:59

Data Download Date/Time: 12/04/2018 14:58

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRANE,RICHARD,B
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: NEWCOMB,WILLIAM
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 10/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: ECK

DATE/TIME OF TEST: 11/21/2018 23:47:46

Diagnostics Check		OK	23:40
Air Blank		0.000	23:40
Control Test		0.080	23:41
Air Blank		0.000	23:41
Subject Sample	# 1	0.198	23:42
Breath Volume		2.292	23:42
Air Blank		0.000	23:43
Air Blank		0.000	23:44
Subject Sample	# 2	0.196	23:46
Breath Volume		2.128	23:46
Air Blank		0.000	23:46
Control Test		0.080	23:47
Air Blank		0.000	23:47
Diagnostics Check		OK	23:47

Data Download Date/Time: 12/04/2018 14:58

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000742

INSTRUMENT REGISTERED TO: VALPARAISO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TONEY,CRAIG,P
BREATH TEST OPERATOR AGENCY: Valparaiso PD
ARRESTING OFFICER: POLLARD,ELIJAH
ARRESTING OFFICER AGENCY: Valparaiso PD
LAST AGENCY INSPECTION: 10/26/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 11/02/2018 23:31:56

Diagnostics Check		OK	23:25
Air Blank		0.000	23:26
Control Test		0.080	23:26
Air Blank		0.000	23:27
Subject Sample	# 1	0.025	23:27
Breath Volume		1.300	23:27
Air Blank		0.000	23:28
Air Blank		0.000	23:30
Subject Sample	# 2	0.027	23:30
Breath Volume		2.703	23:30
Air Blank		0.000	23:31
Control Test		0.080	23:31
Air Blank		0.000	23:31
Diagnostics Check		OK	23:31

Data Download Date/Time: 11/19/2018 15:36

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000742

INSTRUMENT REGISTERED TO: VALPARAISO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIDSON,JEREMIAH,D
BREATH TEST OPERATOR AGENCY: Valparaiso PD
ARRESTING OFFICER: PRICE,ALPHONSO
ARRESTING OFFICER AGENCY: Eglin AFB PD
LAST AGENCY INSPECTION: 10/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: DALE

DATE/TIME OF TEST: 11/12/2018 02:00:54

Diagnostics Check		OK	01:53
Air Blank		0.000	01:54
Control Test		0.079	01:54
Air Blank		0.000	01:55
Subject Sample	# 1	0.000	01:56
Breath Volume		1.691	01:56
Air Blank		0.000	01:57
Air Blank		0.000	01:59
Subject Sample	# 2	0.000	01:59
Breath Volume		2.050	01:59
Air Blank		0.000	02:00
Control Test		0.078	02:00
Air Blank		0.000	02:00
Diagnostics Check		OK	02:00

Data Download Date/Time: 11/19/2018 15:36

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000742

INSTRUMENT REGISTERED TO: VALPARAISO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIDSON,JEREMIAH,D
BREATH TEST OPERATOR AGENCY: Valparaiso PD
ARRESTING OFFICER: PRICE,ALPHONSO
ARRESTING OFFICER AGENCY: Eglin AFB PD
LAST AGENCY INSPECTION: 10/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: SALGADO

DATE/TIME OF TEST: 11/12/2018 02:11:43

Diagnostics Check		OK	02:04
Air Blank		0.000	02:04
Control Test		0.078	02:05
Air Blank		0.000	02:05
Subject Sample	# 1	0.090	02:07
Breath Volume		2.195	02:07
Air Blank		0.000	02:07
Air Blank		0.000	02:09
Subject Sample	# 2	0.085	02:10
Breath Volume		2.011	02:10
Air Blank		0.000	02:10
Control Test		0.080	02:11
Air Blank		0.000	02:11
Diagnostics Check		OK	02:11

Data Download Date/Time: 11/19/2018 15:36

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUDSON,JOSHUA,K
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: COLLINS,CHASE
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 10/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: THALER

DATE/TIME OF TEST: 11/01/2018 02:35:28

Diagnostics Check		OK	02:27
Air Blank		0.000	02:28
Control Test		0.079	02:28
Air Blank		0.000	02:28
Subject Sample	# 1	0.233	02:30
Breath Volume		2.238	02:30
Air Blank		0.000	02:31
Air Blank		0.000	02:33
Subject Sample	# 2	0.246	02:33
Breath Volume		1.464	02:33
Air Blank		0.000	02:34
Control Test		0.078	02:34
Air Blank		0.000	02:35
Diagnostics Check		OK	02:35

Data Download Date/Time: 11/26/2018 07:14

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,AMANDA,L
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: RUBIO GARCIA,ALVARO
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 10/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:08

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 11/03/2018 18:49:42

Diagnostics Check		OK	18:43
Air Blank		0.000	18:44
Control Test		0.080	18:44
Air Blank		0.000	18:44
Subject Sample	# 1	0.000	18:45
Breath Volume		4.343	18:45
Air Blank		0.000	18:45
Air Blank		0.000	18:47
Subject Sample	# 2	0.000	18:48
Breath Volume		4.230	18:48
Air Blank		0.000	18:48
Control Test		0.080	18:49
Air Blank		0.000	18:49
Diagnostics Check		OK	18:49

Data Download Date/Time: 11/26/2018 07:14

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BYRD,BRANDON,C
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: PARLIER,STEVEN
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 10/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: KORTLEVER

DATE/TIME OF TEST: 11/04/2018 01:48:48

Diagnostics Check		OK	01:42
Air Blank		0.000	01:42
Control Test		0.080	01:42
Air Blank		0.000	01:43
Subject Sample	# 1	0.174	01:44
Breath Volume		2.667	01:44
Air Blank		0.000	01:44
Air Blank		0.000	01:46
Subject Sample	# 2	0.174	01:47
Breath Volume		2.738	01:47
Air Blank		0.000	01:47
Control Test		0.079	01:48
Air Blank		0.000	01:48
Diagnostics Check		OK	01:48

Data Download Date/Time: 11/26/2018 07:14

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SALON,DENNIS,M
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: HUAL,MICHAEL
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 10/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:35

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 11/05/2018 06:03:16

Diagnostics Check		OK	05:55
Air Blank		0.000	05:56
Control Test		0.078	05:56
Air Blank		0.000	05:56
Subject Sample	# 1	0.147	05:57
Breath Volume		3.371	05:57
Air Blank		0.000	05:58
Air Blank		0.000	06:00
Subject Sample	# 2	0.144	06:01
Breath Volume		3.476	06:01
Air Blank		0.000	06:02
Control Test		0.077	06:02
Air Blank		0.000	06:03
Diagnostics Check		OK	06:03

Data Download Date/Time: 11/26/2018 07:14

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEEKS,JEREMIAH,S
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: GLASS,JONATHAN
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 10/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: KNIGHT

DATE/TIME OF TEST: 11/11/2018 04:43:47

Diagnostics Check		OK	04:40	
Air Blank		0.000	04:40	
Control Test		0.082	04:41	
Air Blank		0.000	04:41	
Subject Sample	# 1		04:42	Subject Test Refused
Breath Volume		0.000	04:42	
Air Blank		0.000	04:42	
Control Test		0.080	04:43	
Air Blank		0.000	04:43	
Diagnostics Check		OK	04:43	

Data Download Date/Time: 11/26/2018 07:14

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARACLE,MATTHEW,D
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: DUFOUR,NOAH
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 10/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 07:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:20

SUBJECT'S LAST NAME: DEBENEDETTO

DATE/TIME OF TEST: 11/11/2018 07:51:12

Diagnostics Check		OK	07:43
Air Blank		0.000	07:43
Control Test		0.080	07:44
Air Blank		0.000	07:44
Subject Sample	# 1	0.000	07:45
Breath Volume		1.210	07:45
Air Blank		0.000	07:45
Air Blank		0.000	07:47
Subject Sample	# 2	0.000	07:49
Breath Volume		1.734	07:49
Air Blank		0.000	07:50
Control Test		0.080	07:50
Air Blank		0.000	07:51
Diagnostics Check		OK	07:51

Data Download Date/Time: 11/26/2018 07:14

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VO,SEBATIAN,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: VO,SEBASTIAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 13:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:22

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 11/11/2018 13:59:25

Diagnostics Check		OK	13:52
Air Blank		0.000	13:53
Control Test		0.080	13:53
Air Blank		0.000	13:54
Subject Sample	# 1	0.077	13:54
Breath Volume		2.035	13:54
Air Blank		0.000	13:55
Air Blank		0.000	13:56
Subject Sample	# 2	0.070	13:57
Breath Volume		1.613	13:57
Air Blank		0.000	13:58
Control Test		0.080	13:58
Air Blank		0.000	13:59
Diagnostics Check		OK	13:59

Data Download Date/Time: 11/26/2018 07:14

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DANIELS,JAMES,A
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: MERAZ,CHRISTOPHER
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 10/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 14:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:17

SUBJECT'S LAST NAME: DILLARD

DATE/TIME OF TEST: 11/11/2018 14:45:50

Diagnostics Check		OK	14:37
Air Blank		0.000	14:38
Control Test		0.079	14:38
Air Blank		0.000	14:38
Subject Sample	# 1	0.301	14:41
Breath Volume		1.839	14:41
Air Blank		0.000	14:41
Air Blank		0.000	14:43
Subject Sample	# 2	0.305	14:44
Breath Volume		2.527	14:44
Air Blank		0.000	14:44
Control Test		0.078	14:45
Air Blank		0.000	14:45
Diagnostics Check		OK	14:45

Data Download Date/Time: 11/26/2018 07:14

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEEKS,JEREMIAH,S
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: MEEKS,JEREMIAH
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 10/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:23

SUBJECT'S LAST NAME: HIGGINS

DATE/TIME OF TEST: 11/11/2018 22:53:43

Diagnostics Check		OK	22:43
Air Blank		0.000	22:44
Control Test		0.079	22:44
Air Blank		0.000	22:45
Subject Sample	# 1	0.189	22:46
Breath Volume		1.265	22:46
Air Blank		0.000	22:47
Air Blank		0.000	22:48
Subject Sample	# 2	0.179	22:52
Breath Volume		1.425	22:52
Air Blank		0.000	22:52
Control Test		0.077	22:53
Air Blank		0.000	22:53
Diagnostics Check		OK	22:53

Data Download Date/Time: 11/26/2018 07:14

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,HENDRICK,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MARTINEZ,HENDRICK
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:19

SUBJECT'S LAST NAME: MARKS

DATE/TIME OF TEST: 11/16/2018 21:45:13

Diagnostics Check		OK	21:39
Air Blank		0.000	21:39
Control Test		0.080	21:40
Air Blank		0.000	21:40
Subject Sample	# 1	0.000	21:41
Breath Volume		1.574	21:41
Air Blank		0.000	21:41
Air Blank		0.000	21:43
Subject Sample	# 2	0.000	21:43
Breath Volume		1.898	21:43
Air Blank		0.000	21:44
Control Test		0.081	21:44
Air Blank		0.000	21:45
Diagnostics Check		OK	21:45

Data Download Date/Time: 11/26/2018 07:14

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,AMANDA,L
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: RUBIO GARCIA,ALVARO
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 10/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: POLNITZ

DATE/TIME OF TEST: 11/18/2018 00:44:25

Diagnostics Check		OK	00:38
Air Blank		0.000	00:38
Control Test		0.079	00:38
Air Blank		0.000	00:39
Subject Sample	# 1	0.190	00:39
Breath Volume		1.351	00:39
Air Blank		0.000	00:40
Air Blank		0.000	00:42
Subject Sample	# 2	0.189	00:42
Breath Volume		1.199	00:42
Air Blank		0.000	00:43
Control Test		0.079	00:43
Air Blank		0.000	00:44
Diagnostics Check		OK	00:44

Data Download Date/Time: 11/26/2018 07:14

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,AMANDA,L
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: HUAL,MICHAEL
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 10/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: LARA DURAN

DATE/TIME OF TEST: 11/18/2018 03:47:50

Diagnostics Check		OK	03:41
Air Blank		0.000	03:42
Control Test		0.080	03:42
Air Blank		0.000	03:42
Subject Sample	# 1	0.165	03:43
Breath Volume		1.921	03:43
Air Blank		0.000	03:44
Air Blank		0.000	03:45
Subject Sample	# 2	0.172	03:46
Breath Volume		2.378	03:46
Air Blank		0.000	03:46
Control Test		0.080	03:47
Air Blank		0.000	03:47
Diagnostics Check		OK	03:47

Data Download Date/Time: 11/26/2018 07:14

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARY,ADAM,C
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: PARLIER,STEVEN
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 10/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:25

SUBJECT'S LAST NAME: PRYOR

DATE/TIME OF TEST: 11/19/2018 05:52:36

Diagnostics Check		OK	05:46
Air Blank		0.000	05:46
Control Test		0.081	05:46
Air Blank		0.000	05:47
Subject Sample	# 1	0.061	05:47
Breath Volume		1.363	05:47
Air Blank		0.000	05:48
Air Blank		0.000	05:50
Subject Sample	# 2	0.070	05:51
Breath Volume		4.035	05:51
Air Blank		0.000	05:51
Control Test		0.080	05:52
Air Blank		0.000	05:52
Diagnostics Check		OK	05:52

Data Download Date/Time: 11/26/2018 07:15

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DILLASHAW,HUNTER,C
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: DILLASHAW,HUNTER
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 10/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: TURECKY

DATE/TIME OF TEST: 11/21/2018 03:51:21

Diagnostics Check		OK	03:44
Air Blank		0.000	03:45
Control Test		0.078	03:45
Air Blank		0.000	03:46
Subject Sample	# 1	0.151	03:46
Breath Volume		1.457	03:46
Air Blank		0.000	03:47
Air Blank		0.000	03:48
Subject Sample	# 2	0.150	03:49
Breath Volume		1.617	03:49
Air Blank		0.000	03:50
Control Test		0.078	03:50
Air Blank		0.000	03:51
Diagnostics Check		OK	03:51

Data Download Date/Time: 11/26/2018 07:15

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOATWRIGHT,JAMIE,R
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: BOATWRIGHT,JAMIE
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 10/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 11/22/2018 01:35:27

Diagnostics Check		OK	01:29
Air Blank		0.000	01:29
Control Test		0.078	01:29
Air Blank		0.000	01:30
Subject Sample	# 1	0.187	01:30
Breath Volume		3.015	01:30
Air Blank		0.000	01:31
Air Blank		0.000	01:33
Subject Sample	# 2	0.198	01:33
Breath Volume		3.183	01:33
Air Blank		0.000	01:34
Control Test		0.079	01:34
Air Blank		0.000	01:35
Diagnostics Check		OK	01:35

Data Download Date/Time: 11/26/2018 07:15

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUDSON,JOSHUA,K
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: MALDEN,BIANCA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 10/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 11/25/2018 01:56:56

Diagnostics Check		OK	01:50
Air Blank		0.000	01:50
Control Test		0.077	01:51
Air Blank		0.000	01:51
Subject Sample	# 1	0.176	01:52
Breath Volume		2.371	01:52
Air Blank		0.000	01:52
Air Blank		0.000	01:54
Subject Sample	# 2	0.180	01:55
Breath Volume		2.578	01:55
Air Blank		0.000	01:56
Control Test		0.077	01:56
Air Blank		0.000	01:56
Diagnostics Check		OK	01:56

Data Download Date/Time: 11/26/2018 07:15

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TASCHETTA,JOSEPH,S
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: BUZBEE,HEATHER
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 11/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: HOLT

DATE/TIME OF TEST: 11/27/2018 00:34:15

Diagnostics Check		OK	00:19	
Air Blank		0.000	00:20	
Control Test		0.079	00:20	
Air Blank		0.000	00:20	
Subject Sample	# 1	0.180	00:21	
Breath Volume		2.566	00:21	
Air Blank		0.000	00:22	
Air Blank		0.000	00:24	
Subject Sample	# 2		00:27	No Sample Provided
Breath Volume		0.000	00:27	
Air Blank		0.000	00:27	
Air Blank		0.000	00:29	
Subject Sample	# 3		00:32	No Sample Provided
Breath Volume		0.000	00:32	
Air Blank		0.000	00:33	
Control Test		0.077	00:33	
Air Blank		0.000	00:34	
Diagnostics Check		OK	00:34	

Data Download Date/Time: 12/19/2018 08:28

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TASCHETTA,JOSEPH,S
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: BUZBEE,HEATHER
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 11/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: HOLT

DATE/TIME OF TEST: 11/27/2018 01:03:56

Diagnostics Check		OK	00:57
Air Blank		0.000	00:57
Control Test		0.079	00:57
Air Blank		0.000	00:58
Subject Sample	# 1	0.158	00:59
Breath Volume		1.519	00:59
Air Blank		0.000	00:59
Air Blank		0.000	01:01
Subject Sample	# 2	0.167	01:02
Breath Volume		2.390	01:02
Air Blank		0.000	01:03
Control Test		0.079	01:03
Air Blank		0.000	01:03
Diagnostics Check		OK	01:03

Data Download Date/Time: 12/19/2018 08:28

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEEKS,JEREMIAH,S
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: MEEKS,JEREMIAH
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 11/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 11/27/2018 03:10:13

Diagnostics Check		OK	03:03
Air Blank		0.000	03:04
Control Test		0.079	03:04
Air Blank		0.000	03:05
Subject Sample	# 1	0.182	03:05
Breath Volume		3.074	03:05
Air Blank		0.000	03:06
Air Blank		0.000	03:08
Subject Sample	# 2	0.185	03:08
Breath Volume		3.324	03:08
Air Blank		0.000	03:09
Control Test		0.078	03:09
Air Blank		0.000	03:10
Diagnostics Check		OK	03:10

Data Download Date/Time: 12/19/2018 08:28

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WORFEL,KYLE,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WORFEL,KYLE
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 875088
DRY GAS EXPIRY DATE: 12/12/2020
VIOLATION CODE: DUI
ARREST TIME: 15:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: OKIEN

DATE/TIME OF TEST: 11/01/2018 16:30:45

Diagnostics Check		OK	16:21
Air Blank		0.000	16:21
Control Test		0.079	16:21
Air Blank		0.000	16:22
Subject Sample	# 1	0.357	16:22
Breath Volume		2.820	16:22
Air Blank		0.000	16:23
Air Blank		0.000	16:25
Subject Sample	# 2	0.313	16:25
Breath Volume		1.414	16:25
Air Blank		0.000	16:26
Air Blank		0.000	16:28
Subject Sample	# 3	0.353	16:29
Breath Volume		1.625	16:29
Air Blank		0.000	16:29
Control Test		0.079	16:30
Air Blank		0.000	16:30
Diagnostics Check		OK	16:30

Data Download Date/Time: 11/16/2018 15:18

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,MICHAEL,T
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: THOMAS,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 875088
DRY GAS EXPIRY DATE: 12/12/2020
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: MACK

DATE/TIME OF TEST: 11/02/2018 03:01:19

Diagnostics Check		OK	02:53
Air Blank		0.000	02:54
Control Test		0.080	02:54
Air Blank		0.000	02:55
Subject Sample	# 1	0.155	02:56
Breath Volume		1.750	02:56
Air Blank		0.000	02:57
Air Blank		0.000	02:58
Subject Sample	# 2	0.172	02:59
Breath Volume		1.437	02:59
Air Blank		0.000	03:00
Control Test		0.077	03:00
Air Blank		0.000	03:01
Diagnostics Check		OK	03:01

Data Download Date/Time: 11/16/2018 15:18

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELRUSSO,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: DELRUSSO,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 875088
DRY GAS EXPIRY DATE: 12/12/2020
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: BRYAN

DATE/TIME OF TEST: 11/02/2018 23:18:42

Diagnostics Check		OK	23:11
Air Blank		0.000	23:11
Control Test		0.080	23:12
Air Blank		0.000	23:12
Subject Sample	# 1	0.213	23:13
Breath Volume		1.347	23:13
Air Blank		0.000	23:14
Air Blank		0.000	23:16
Subject Sample	# 2	0.207	23:17
Breath Volume		1.109	23:17
Air Blank		0.000	23:17
Control Test		0.078	23:18
Air Blank		0.000	23:18
Diagnostics Check		OK	23:18

Data Download Date/Time: 11/16/2018 15:18

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINAN,CHRISTIAN,
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: LINAN,CHRISTIAN
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 10/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 875088
DRY GAS EXPIRY DATE: 12/12/2020
VIOLATION CODE: DUI
ARREST TIME: 09:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:15

SUBJECT'S LAST NAME: COSTALES

DATE/TIME OF TEST: 11/06/2018 10:39:51

Diagnostics Check		OK	10:36	
Air Blank		0.000	10:37	
Control Test		0.080	10:37	
Air Blank		0.000	10:38	
Subject Sample	# 1		10:38	Subject Test Refused
Breath Volume		0.000	10:38	
Air Blank		0.000	10:38	
Control Test		0.079	10:39	
Air Blank		0.000	10:39	
Diagnostics Check		OK	10:39	

Data Download Date/Time: 11/16/2018 15:18

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REKULLY,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: TRUMAN,RICHARD
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 10/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 875088
DRY GAS EXPIRY DATE: 12/12/2020
VIOLATION CODE: DUI
ARREST TIME: 09:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:52

SUBJECT'S LAST NAME: ZASTROW

DATE/TIME OF TEST: 11/07/2018 11:28:26

Diagnostics Check		OK	11:21
Air Blank		0.000	11:21
Control Test		0.080	11:22
Air Blank		0.000	11:22
Subject Sample	# 1	0.000	11:24
Breath Volume		2.734	11:24
Air Blank		0.000	11:24
Air Blank		0.000	11:26
Subject Sample	# 2	0.000	11:27
Breath Volume		2.343	11:27
Air Blank		0.000	11:27
Control Test		0.078	11:27
Air Blank		0.000	11:28
Diagnostics Check		OK	11:28

Data Download Date/Time: 11/16/2018 15:18

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: UMANA,RAUL,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: UMANA,RAUL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 875088
DRY GAS EXPIRY DATE: 12/12/2020
VIOLATION CODE: DUI
ARREST TIME: 20:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: PHIFIELD

DATE/TIME OF TEST: 11/10/2018 20:54:06

Diagnostics Check		OK	20:51	
Air Blank		0.000	20:51	
Control Test		0.079	20:51	
Air Blank		0.000	20:52	
Subject Sample	# 1		20:52	Subject Test Refused
Breath Volume		0.000	20:52	
Air Blank		0.000	20:53	
Control Test		0.079	20:53	
Air Blank		0.000	20:54	
Diagnostics Check		OK	20:54	

Data Download Date/Time: 11/16/2018 15:18

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: UMANA,RAUL,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: UMANA,RAUL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 875088
DRY GAS EXPIRY DATE: 12/12/2020
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: ZAVADIL

DATE/TIME OF TEST: 11/11/2018 02:01:53

Diagnostics Check		OK	01:55
Air Blank		0.000	01:55
Control Test		0.080	01:56
Air Blank		0.000	01:56
Subject Sample	# 1	0.228	01:57
Breath Volume		2.300	01:57
Air Blank		0.000	01:57
Air Blank		0.000	01:59
Subject Sample	# 2	0.237	02:00
Breath Volume		2.382	02:00
Air Blank		0.000	02:00
Control Test		0.078	02:01
Air Blank		0.000	02:01
Diagnostics Check		OK	02:01

Data Download Date/Time: 11/16/2018 15:18

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAFFERTY,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: RAFFERTY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 10/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 875088
DRY GAS EXPIRY DATE: 12/12/2020
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: AMEND

DATE/TIME OF TEST: 11/11/2018 03:58:58

Diagnostics Check		OK	03:52
Air Blank		0.000	03:53
Control Test		0.080	03:53
Air Blank		0.000	03:53
Subject Sample	# 1	0.096	03:54
Breath Volume		2.808	03:54
Air Blank		0.000	03:55
Air Blank		0.000	03:56
Subject Sample	# 2	0.093	03:57
Breath Volume		2.578	03:57
Air Blank		0.000	03:58
Control Test		0.077	03:58
Air Blank		0.000	03:58
Diagnostics Check		OK	03:58

Data Download Date/Time: 11/16/2018 15:18

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILLIARD,JEFFERY,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: HILLIARD,JEFFERY
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 875088
DRY GAS EXPIRY DATE: 12/12/2020
VIOLATION CODE: DUI
ARREST TIME: 14:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:15

SUBJECT'S LAST NAME: PARSONS

DATE/TIME OF TEST: 11/11/2018 15:42:40

Diagnostics Check		OK	15:36
Air Blank		0.000	15:36
Control Test		0.080	15:36
Air Blank		0.000	15:37
Subject Sample	# 1	0.119	15:38
Breath Volume		1.765	15:38
Air Blank		0.000	15:38
Air Blank		0.000	15:40
Subject Sample	# 2	0.113	15:41
Breath Volume		1.890	15:41
Air Blank		0.000	15:41
Control Test		0.078	15:42
Air Blank		0.000	15:42
Diagnostics Check		OK	15:42

Data Download Date/Time: 11/16/2018 15:18

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRINGTON,ALEX,M
BREATH TEST OPERATOR AGENCY: Belleview PD
ARRESTING OFFICER: HARRINGTON,ALEX
ARRESTING OFFICER AGENCY: Belleview PD
LAST AGENCY INSPECTION: 10/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 875088
DRY GAS EXPIRY DATE: 12/12/2020
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: FUENTES RIVERA

DATE/TIME OF TEST: 11/13/2018 01:15:47

Diagnostics Check		OK	01:12	
Air Blank		0.000	01:12	
Control Test		0.080	01:13	
Air Blank		0.000	01:13	
Subject Sample	# 1		01:14	Subject Test Refused
Breath Volume		0.000	01:14	
Air Blank		0.000	01:14	
Control Test		0.080	01:15	
Air Blank		0.000	01:15	
Diagnostics Check		OK	01:15	

Data Download Date/Time: 11/16/2018 15:18

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LACOGNATA,CHRISTINA,T
BREATH TEST OPERATOR AGENCY: Belleview PD
ARRESTING OFFICER: LACOGNATA,CHRISTINA
ARRESTING OFFICER AGENCY: Belleview PD
LAST AGENCY INSPECTION: 10/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 875088
DRY GAS EXPIRY DATE: 12/12/2020
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: DIETZEL

DATE/TIME OF TEST: 11/13/2018 03:30:32

Diagnostics Check		OK	03:23
Air Blank		0.000	03:23
Control Test		0.080	03:23
Air Blank		0.000	03:24
Subject Sample	# 1	0.230	03:25
Breath Volume		2.074	03:25
Air Blank		0.000	03:26
Air Blank		0.000	03:28
Subject Sample	# 2	0.223	03:28
Breath Volume		3.949	03:28
Air Blank		0.000	03:29
Control Test		0.079	03:29
Air Blank		0.000	03:30
Diagnostics Check		OK	03:30

Data Download Date/Time: 11/16/2018 15:18

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STILLION,JUSTIN,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: STILLION,JUSTIN
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 10/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 875088
DRY GAS EXPIRY DATE: 12/12/2020
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: PEDRICK

DATE/TIME OF TEST: 11/14/2018 23:41:58

Diagnostics Check		OK	23:34	
Air Blank		0.000	23:35	
Control Test		0.080	23:35	
Air Blank		0.000	23:36	
Subject Sample	# 1	0.112	23:36	
Breath Volume		2.910	23:36	
Air Blank		0.000	23:38	
Air Blank		0.000	23:39	
Subject Sample	# 2	0.111	23:40	
Breath Volume		3.253	23:40	
Air Blank		0.000	23:41	
Control Test		0.074	23:41	Control Outside Tolerance
Air Blank		0.000	23:41	

Data Download Date/Time: 11/16/2018 15:18

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOLDSMITH,WILLIE,E
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MCNEILL,CONNIE
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 875088
DRY GAS EXPIRY DATE: 12/12/2020
VIOLATION CODE: DUI
ARREST TIME: 10:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:00

SUBJECT'S LAST NAME: G432885632920

DATE/TIME OF TEST: 11/15/2018 11:23:07

Diagnostics Check	OK	11:22	
Air Blank		11:22	Ambient Fail
Air Blank	0.000	11:23	

Data Download Date/Time: 11/16/2018 15:18

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCNEILL,CONNIE,F
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MCNEILL,CONNIE
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/29/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 875088
DRY GAS EXPIRY DATE: 12/12/2020
VIOLATION CODE: DUI
ARREST TIME: 10:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:00

SUBJECT'S LAST NAME: GOLDSMITH

DATE/TIME OF TEST: 11/15/2018 11:30:20

Diagnostics Check		OK	11:27	
Air Blank		0.000	11:27	
Control Test		0.076	11:27	
Air Blank		0.000	11:28	
Subject Sample	# 1	0.146	11:29	
Breath Volume		2.582	11:29	
Air Blank			11:29	Purge Fail
Air Blank		0.000	11:30	

Data Download Date/Time: 11/16/2018 15:18

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCNEILL,CONNIE,F
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MCNEILL,CONNIE
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 875088
DRY GAS EXPIRY DATE: 12/12/2020
VIOLATION CODE: DUI
ARREST TIME: 10:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:00

SUBJECT'S LAST NAME: GOLSMITH

DATE/TIME OF TEST: 11/15/2018 11:39:12

Diagnostics Check		OK	11:33	
Air Blank		0.000	11:33	
Control Test		0.078	11:33	
Air Blank		0.000	11:34	
Subject Sample	# 1	0.144	11:35	
Breath Volume		1.882	11:35	
Air Blank		0.000	11:36	
Air Blank		0.000	11:37	
Subject Sample	# 2		11:38	RFI Detect
Breath Volume		0.000	11:38	
Air Blank			11:38	RFI Detect
Air Blank		0.000	11:39	

Data Download Date/Time: 11/16/2018 15:18

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCNEILL,CONNIE,F
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MCNEILL,CONNIE
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 875088
DRY GAS EXPIRY DATE: 12/12/2020
VIOLATION CODE: DUI
ARREST TIME: 10:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:00

SUBJECT'S LAST NAME: GOLDSMITH

DATE/TIME OF TEST: 11/15/2018 11:45:12

Diagnostics Check		OK	11:41	
Air Blank		0.000	11:42	
Control Test		0.079	11:42	
Air Blank		0.000	11:43	
Subject Sample	# 1		11:43	RFI Detect
Breath Volume		0.000	11:43	
Air Blank		0.000	11:44	
Control Test		0.078	11:44	
Air Blank		0.000	11:45	
Diagnostics Check		OK	11:45	

Data Download Date/Time: 11/16/2018 15:18

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCNEILL,CONNIE,F
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MCNEILL,CONNIE
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 875088
DRY GAS EXPIRY DATE: 12/12/2020
VIOLATION CODE: DUI
ARREST TIME: 10:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:00

SUBJECT'S LAST NAME: GOLDSMITH

DATE/TIME OF TEST: 11/15/2018 11:51:59

Diagnostics Check		OK	11:47	
Air Blank		0.000	11:48	
Control Test		0.078	11:48	
Air Blank		0.000	11:49	
Subject Sample	# 1	0.131	11:49	
Breath Volume		1.164	11:49	
Air Blank			11:50	RFI Detect
Air Blank		0.000	11:50	
Control Test		0.078	11:51	
Air Blank		0.000	11:51	
Diagnostics Check		OK	11:51	

Data Download Date/Time: 11/16/2018 15:18

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,MICHAEL,T
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: THOMAS,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 875088
DRY GAS EXPIRY DATE: 12/12/2020
VIOLATION CODE: DUI
ARREST TIME: 06:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:20

SUBJECT'S LAST NAME: BARNES

DATE/TIME OF TEST: 11/15/2018 18:35:18

Diagnostics Check		OK	18:28
Air Blank		0.000	18:29
Control Test		0.080	18:29
Air Blank		0.000	18:30
Subject Sample	# 1	0.179	18:30
Breath Volume		2.125	18:30
Air Blank		0.000	18:31
Air Blank		0.000	18:33
Subject Sample	# 2	0.182	18:33
Breath Volume		2.484	18:33
Air Blank		0.000	18:34
Control Test		0.076	18:34
Air Blank		0.000	18:35
Diagnostics Check		OK	18:35

Data Download Date/Time: 11/16/2018 15:18

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOOREHEAD,DAVID,L
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: CABRERA,NICKESHA
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 11/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 875088
DRY GAS EXPIRY DATE: 12/12/2020
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: VEALE

DATE/TIME OF TEST: 11/17/2018 03:09:49

Diagnostics Check		OK	03:03
Air Blank		0.000	03:03
Control Test		0.080	03:03
Air Blank		0.000	03:04
Subject Sample	# 1	0.239	03:05
Breath Volume		2.980	03:05
Air Blank		0.000	03:05
Air Blank		0.000	03:07
Subject Sample	# 2	0.240	03:08
Breath Volume		2.605	03:08
Air Blank		0.000	03:08
Control Test		0.079	03:09
Air Blank		0.000	03:09
Diagnostics Check		OK	03:09

Data Download Date/Time: 12/21/2018 13:09

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BALL,CODY,W
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BALL,CODY
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 11/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 875088
DRY GAS EXPIRY DATE: 12/12/2020
VIOLATION CODE: DUI
ARREST TIME: 18:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:49

SUBJECT'S LAST NAME: DAVISON

DATE/TIME OF TEST: 11/17/2018 20:23:22

Diagnostics Check		OK	20:14
Air Blank		0.000	20:14
Control Test		0.080	20:14
Air Blank		0.000	20:15
Subject Sample	# 1	0.108	20:16
Breath Volume		1.656	20:16
Air Blank		0.000	20:17
Air Blank		0.000	20:19
Subject Sample	# 2	0.109	20:21
Breath Volume		2.007	20:21
Air Blank		0.000	20:22
Control Test		0.079	20:22
Air Blank		0.000	20:23
Diagnostics Check		OK	20:23

Data Download Date/Time: 12/21/2018 13:09

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTIGLIONE,S,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: CASTIGLIONE,S
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 11/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 875088
DRY GAS EXPIRY DATE: 12/12/2020
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: JIMINEZ SANTIAGO

DATE/TIME OF TEST: 11/18/2018 03:25:17

Diagnostics Check		OK	03:15
Air Blank		0.000	03:16
Control Test		0.080	03:16
Air Blank		0.000	03:17
Subject Sample	# 1	0.128	03:19
Breath Volume		1.371	03:19
Air Blank		0.000	03:20
Air Blank		0.000	03:22
Subject Sample	# 2	0.133	03:23
Breath Volume		1.867	03:23
Air Blank		0.000	03:24
Control Test		0.077	03:24
Air Blank		0.000	03:25
Diagnostics Check		OK	03:25

Data Download Date/Time: 12/21/2018 13:09

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYS,MATTHEW,W
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: CASTIGLIONE,S
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 11/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 875088
DRY GAS EXPIRY DATE: 12/12/2020
VIOLATION CODE: DUI
ARREST TIME: 20:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:17

SUBJECT'S LAST NAME: GIDEON

DATE/TIME OF TEST: 11/18/2018 21:44:32

Diagnostics Check		OK	21:38
Air Blank		0.000	21:38
Control Test		0.079	21:38
Air Blank		0.000	21:39
Subject Sample	# 1	0.238	21:39
Breath Volume		3.128	21:39
Air Blank		0.000	21:40
Air Blank		0.000	21:42
Subject Sample	# 2	0.223	21:42
Breath Volume		1.464	21:42
Air Blank		0.000	21:43
Control Test		0.079	21:43
Air Blank		0.000	21:44
Diagnostics Check		OK	21:44

Data Download Date/Time: 12/21/2018 13:09

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LETSON,REBECCA,G
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: LETSON,REBECCA
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 11/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 875088
DRY GAS EXPIRY DATE: 12/12/2020
VIOLATION CODE: DUI
ARREST TIME: 19:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:47

SUBJECT'S LAST NAME: CRUMMEY

DATE/TIME OF TEST: 11/19/2018 20:18:19

Diagnostics Check		OK	20:12
Air Blank		0.000	20:12
Control Test		0.081	20:12
Air Blank		0.000	20:13
Subject Sample	# 1	0.000	20:14
Breath Volume		1.851	20:14
Air Blank		0.000	20:14
Air Blank		0.000	20:16
Subject Sample	# 2	0.000	20:16
Breath Volume		1.640	20:16
Air Blank		0.000	20:17
Control Test		0.080	20:17
Air Blank		0.000	20:18
Diagnostics Check		OK	20:18

Data Download Date/Time: 12/21/2018 13:10

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERSON,KIP,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: PETERSON,KIP
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 11/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 875088
DRY GAS EXPIRY DATE: 12/12/2020
VIOLATION CODE: DUI
ARREST TIME: 11:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:48

SUBJECT'S LAST NAME: DAMMS

DATE/TIME OF TEST: 11/21/2018 14:29:15

Diagnostics Check		OK	14:22
Air Blank		0.000	14:23
Control Test		0.080	14:23
Air Blank		0.000	14:24
Subject Sample	# 1	0.000	14:24
Breath Volume		1.109	14:24
Air Blank		0.000	14:25
Air Blank		0.000	14:27
Subject Sample	# 2	0.000	14:27
Breath Volume		1.441	14:27
Air Blank		0.000	14:28
Control Test		0.078	14:28
Air Blank		0.000	14:29
Diagnostics Check		OK	14:29

Data Download Date/Time: 12/21/2018 13:10

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ATKINSON,BLEYS,R
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: DEMARCO,RICHARD
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 11/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 875088
DRY GAS EXPIRY DATE: 12/20/2020
VIOLATION CODE: DUI
ARREST TIME: 18:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:37

SUBJECT'S LAST NAME: MOBLEY

DATE/TIME OF TEST: 11/21/2018 21:11:50

Diagnostics Check		OK	20:59	
Air Blank		0.000	21:00	
Control Test		0.080	21:00	
Air Blank		0.000	21:01	
Subject Sample	# 1	0.042	21:04	Volume Not Met
Breath Volume		0.242	21:04	
Air Blank		0.000	21:04	
Air Blank		0.000	21:06	
Subject Sample	# 2	0.042	21:09	Volume Not Met
Breath Volume		0.285	21:09	
Air Blank		0.000	21:10	
Control Test		0.079	21:11	
Air Blank		0.000	21:11	
Diagnostics Check		OK	21:11	

Data Download Date/Time: 12/21/2018 13:10

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ATKINSON,BLEYS,R
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: DEMARCO,RICHARD
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 11/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 875088
DRY GAS EXPIRY DATE: 12/20/2020
VIOLATION CODE: DUI
ARREST TIME: 18:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:37

SUBJECT'S LAST NAME: MOBLEY

DATE/TIME OF TEST: 11/21/2018 21:29:48

Diagnostics Check		OK	21:22
Air Blank		0.000	21:22
Control Test		0.078	21:23
Air Blank		0.000	21:23
Subject Sample	# 1	0.040	21:25
Breath Volume		2.417	21:25
Air Blank		0.000	21:25
Air Blank		0.000	21:27
Subject Sample	# 2	0.035	21:27
Breath Volume		1.433	21:27
Air Blank		0.000	21:28
Control Test		0.077	21:29
Air Blank		0.000	21:29
Diagnostics Check		OK	21:29

Data Download Date/Time: 12/21/2018 13:10

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANCHESTER,MILTON,S
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: ROSACI,NEIL
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 11/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 875088
DRY GAS EXPIRY DATE: 12/12/2020
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: WURSTER

DATE/TIME OF TEST: 11/28/2018 05:32:55

Diagnostics Check		OK	05:30	
Air Blank		0.000	05:30	
Control Test		0.081	05:30	
Air Blank		0.000	05:31	
Subject Sample	# 1		05:31	Subject Test Refused
Breath Volume		0.000	05:31	
Air Blank		0.000	05:31	
Control Test		0.081	05:32	
Air Blank		0.000	05:32	
Diagnostics Check		OK	05:32	

Data Download Date/Time: 12/21/2018 13:10

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAFFERTY,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: RAFFERTY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 11/16/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 875088
DRY GAS EXPIRY DATE: 12/12/2020
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: IMES

DATE/TIME OF TEST: 11/30/2018 00:25:10

Diagnostics Check		OK	00:19
Air Blank		0.000	00:19
Control Test		0.082	00:19
Air Blank		0.000	00:20
Subject Sample	# 1	0.000	00:20
Breath Volume		1.617	00:20
Air Blank		0.000	00:21
Air Blank		0.000	00:23
Subject Sample	# 2	0.000	00:23
Breath Volume		2.746	00:23
Air Blank		0.000	00:24
Control Test		0.078	00:24
Air Blank		0.000	00:25
Diagnostics Check		OK	00:25

Data Download Date/Time: 12/21/2018 13:10

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,D,E
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: WRIGHT,D
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 11/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 875088
DRY GAS EXPIRY DATE: 12/12/2020
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 11/30/2018 01:07:53

Diagnostics Check		OK	01:01
Air Blank		0.000	01:01
Control Test		0.079	01:02
Air Blank		0.000	01:02
Subject Sample	# 1	0.112	01:03
Breath Volume		1.964	01:03
Air Blank		0.000	01:03
Air Blank		0.000	01:05
Subject Sample	# 2	0.115	01:06
Breath Volume		2.050	01:06
Air Blank		0.000	01:06
Control Test		0.080	01:07
Air Blank		0.000	01:07
Diagnostics Check		OK	01:07

Data Download Date/Time: 12/21/2018 13:10

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOOREHEAD,DAVID,L
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: MOOREHEAD,DAVID
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 11/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 875088
DRY GAS EXPIRY DATE: 12/12/2020
VIOLATION CODE: DUI
ARREST TIME: 21:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:32

SUBJECT'S LAST NAME: MARES

DATE/TIME OF TEST: 11/30/2018 21:59:54

Diagnostics Check		OK	21:52
Air Blank		0.000	21:52
Control Test		0.080	21:53
Air Blank		0.000	21:53
Subject Sample	# 1	0.201	21:54
Breath Volume		1.761	21:54
Air Blank		0.000	21:55
Air Blank		0.000	21:57
Subject Sample	# 2	0.203	21:58
Breath Volume		1.949	21:58
Air Blank		0.000	21:58
Control Test		0.078	21:59
Air Blank		0.000	21:59
Diagnostics Check		OK	21:59

Data Download Date/Time: 12/21/2018 13:10

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWELL,KEENAN,B
BREATH TEST OPERATOR AGENCY: Marianna PD
ARRESTING OFFICER: ARROYO,TRISTON
ARRESTING OFFICER AGENCY: Marianna PD
LAST AGENCY INSPECTION: 11/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: FREEMAN

DATE/TIME OF TEST: 11/15/2018 02:33:41

Diagnostics Check		OK	02:27
Air Blank		0.000	02:27
Control Test		0.082	02:28
Air Blank		0.000	02:28
Subject Sample	# 1	0.000	02:29
Breath Volume		1.675	02:29
Air Blank		0.000	02:29
Air Blank		0.000	02:31
Subject Sample	# 2	0.000	02:32
Breath Volume		4.070	02:32
Air Blank		0.000	02:32
Control Test		0.080	02:33
Air Blank		0.000	02:33
Diagnostics Check		OK	02:33

Data Download Date/Time: 12/01/2018 03:34

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRISP,JESSICA,M
BREATH TEST OPERATOR AGENCY: Jackson County SO
ARRESTING OFFICER: RAFFERTY,JOSHUA
ARRESTING OFFICER AGENCY: Jackson County SO
LAST AGENCY INSPECTION: 11/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: DAVIS JR

DATE/TIME OF TEST: 11/22/2018 02:56:49

Diagnostics Check		OK	02:47
Air Blank		0.000	02:48
Control Test		0.081	02:48
Air Blank		0.000	02:49
Subject Sample	# 1	0.148	02:52
Breath Volume		2.191	02:52
Air Blank		0.000	02:53
Air Blank		0.000	02:54
Subject Sample	# 2	0.148	02:55
Breath Volume		1.230	02:55
Air Blank		0.000	02:55
Control Test		0.078	02:56
Air Blank		0.000	02:56
Diagnostics Check		OK	02:56

Data Download Date/Time: 12/01/2018 03:34

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROEDEL,ANDREW,D
BREATH TEST OPERATOR AGENCY: Jackson County SO
ARRESTING OFFICER: HAYES,JAMES
ARRESTING OFFICER AGENCY: Graceville PD
LAST AGENCY INSPECTION: 11/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 19:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: UBER

DATE/TIME OF TEST: 11/23/2018 22:12:35

Diagnostics Check		OK	22:06
Air Blank		0.000	22:06
Control Test		0.082	22:07
Air Blank		0.000	22:07
Subject Sample	# 1	0.000	22:08
Breath Volume		2.066	22:08
Air Blank		0.000	22:08
Air Blank		0.000	22:10
Subject Sample	# 2	0.000	22:11
Breath Volume		2.074	22:11
Air Blank		0.000	22:11
Control Test		0.082	22:11
Air Blank		0.000	22:12
Diagnostics Check		OK	22:12

Data Download Date/Time: 12/01/2018 03:34

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SABO,JESSICA,L
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SABO,JESSICA
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: CHAVEZ

DATE/TIME OF TEST: 11/01/2018 04:58:07

Diagnostics Check		OK	04:51
Air Blank		0.000	04:52
Control Test		0.080	04:52
Air Blank		0.000	04:52
Subject Sample	# 1	0.153	04:53
Breath Volume		5.035	04:53
Air Blank		0.000	04:54
Air Blank		0.000	04:56
Subject Sample	# 2	0.150	04:56
Breath Volume		5.101	04:56
Air Blank		0.000	04:57
Control Test		0.080	04:57
Air Blank		0.000	04:58
Diagnostics Check		OK	04:58

Data Download Date/Time: 11/27/2018 10:08

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,CHRISTOPHER,
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GUYTON,JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 10/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 11/03/2018 04:39:12

Diagnostics Check		OK	04:32
Air Blank		0.000	04:32
Control Test		0.079	04:33
Air Blank		0.000	04:33
Subject Sample	# 1	0.134	04:34
Breath Volume		1.839	04:34
Air Blank		0.000	04:35
Air Blank		0.000	04:37
Subject Sample	# 2	0.124	04:37
Breath Volume		1.785	04:37
Air Blank		0.000	04:38
Control Test		0.079	04:38
Air Blank		0.000	04:39
Diagnostics Check		OK	04:39

Data Download Date/Time: 11/27/2018 10:08

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRICKER,JEFFREY,S
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: FRICKER,JEFFREY
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 10/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 11/08/2018 04:50:22

Diagnostics Check		OK	04:44
Air Blank		0.000	04:44
Control Test		0.080	04:44
Air Blank		0.000	04:45
Subject Sample	# 1	0.160	04:45
Breath Volume		2.957	04:45
Air Blank		0.000	04:46
Air Blank		0.000	04:48
Subject Sample	# 2	0.157	04:48
Breath Volume		2.847	04:48
Air Blank		0.000	04:49
Control Test		0.079	04:49
Air Blank		0.000	04:50
Diagnostics Check		OK	04:50

Data Download Date/Time: 11/27/2018 10:08

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELCH,CHAD,J
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: WELCH,CHAD
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 10/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: GROVES

DATE/TIME OF TEST: 11/09/2018 03:50:56

Diagnostics Check		OK	03:44
Air Blank		0.000	03:44
Control Test		0.080	03:44
Air Blank		0.000	03:45
Subject Sample	# 1	0.174	03:46
Breath Volume		1.871	03:46
Air Blank		0.000	03:47
Air Blank		0.000	03:48
Subject Sample	# 2	0.173	03:49
Breath Volume		1.601	03:49
Air Blank		0.000	03:50
Control Test		0.080	03:50
Air Blank		0.000	03:50
Diagnostics Check		OK	03:50

Data Download Date/Time: 11/27/2018 10:08

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUTSCHER,REBECCA,L
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: BUTSCHER,REBECCA
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 10/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:02

SUBJECT'S LAST NAME: CASTILLO RIVERA

DATE/TIME OF TEST: 11/09/2018 21:36:02

Diagnostics Check		OK	21:28
Air Blank		0.000	21:29
Control Test		0.079	21:29
Air Blank		0.000	21:30
Subject Sample	# 1	0.159	21:31
Breath Volume		2.121	21:31
Air Blank		0.000	21:32
Air Blank		0.000	21:33
Subject Sample	# 2	0.158	21:34
Breath Volume		2.324	21:34
Air Blank		0.000	21:35
Control Test		0.078	21:35
Air Blank		0.000	21:35
Diagnostics Check		OK	21:36

Data Download Date/Time: 11/27/2018 10:08

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURINGTON,PATRICK,A
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: PURINGTON,PATRICK
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 10/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: HODGE

DATE/TIME OF TEST: 11/15/2018 05:37:19

Diagnostics Check		OK	05:30
Air Blank		0.000	05:30
Control Test		0.079	05:31
Air Blank		0.000	05:31
Subject Sample	# 1	0.183	05:32
Breath Volume		2.414	05:32
Air Blank		0.000	05:33
Air Blank		0.000	05:35
Subject Sample	# 2	0.182	05:35
Breath Volume		2.230	05:35
Air Blank		0.000	05:36
Control Test		0.078	05:36
Air Blank		0.000	05:37
Diagnostics Check		OK	05:37

Data Download Date/Time: 11/27/2018 10:08

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURPHY,JAVON,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MURPHY,JAVON
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 11/16/2018 23:57:45

Diagnostics Check		OK	23:50
Air Blank		0.000	23:51
Control Test		0.080	23:51
Air Blank		0.000	23:52
Subject Sample	# 1	0.000	23:53
Breath Volume		2.734	23:53
Air Blank		0.000	23:53
Air Blank		0.000	23:55
Subject Sample	# 2	0.000	23:56
Breath Volume		2.996	23:56
Air Blank		0.000	23:56
Control Test		0.080	23:57
Air Blank		0.000	23:57
Diagnostics Check		OK	23:57

Data Download Date/Time: 11/27/2018 10:08

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBLES,MIGUEL,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: ROBLES,MIGUEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: ROWLAND

DATE/TIME OF TEST: 11/18/2018 00:25:29

Diagnostics Check		OK	00:19
Air Blank		0.000	00:19
Control Test		0.080	00:19
Air Blank		0.000	00:20
Subject Sample	# 1	0.171	00:20
Breath Volume		2.753	00:20
Air Blank		0.000	00:21
Air Blank		0.000	00:23
Subject Sample	# 2	0.181	00:23
Breath Volume		4.550	00:23
Air Blank		0.000	00:24
Control Test		0.079	00:24
Air Blank		0.000	00:25
Diagnostics Check		OK	00:25

Data Download Date/Time: 11/27/2018 10:08

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LITZKOW,CARLTON,D
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: LITZKOW,CARLTON
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 10/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 11/18/2018 05:46:14

Diagnostics Check		OK	05:39
Air Blank		0.000	05:40
Control Test		0.079	05:40
Air Blank		0.000	05:40
Subject Sample	# 1	0.010	05:41
Breath Volume		3.339	05:41
Air Blank		0.000	05:42
Air Blank		0.000	05:44
Subject Sample	# 2	0.010	05:44
Breath Volume		3.589	05:44
Air Blank		0.000	05:45
Control Test		0.080	05:45
Air Blank		0.000	05:46
Diagnostics Check		OK	05:46

Data Download Date/Time: 11/27/2018 10:08

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNS,WILLIAM,A
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: STANFIELD,THOMAS
ARRESTING OFFICER AGENCY: Alachua PD
LAST AGENCY INSPECTION: 10/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 14:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:34

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 11/18/2018 16:04:10

Diagnostics Check		OK	15:57
Air Blank		0.000	15:57
Control Test		0.079	15:57
Air Blank		0.000	15:58
Subject Sample	# 1	0.209	15:59
Breath Volume		2.289	15:59
Air Blank		0.000	16:00
Air Blank		0.000	16:02
Subject Sample	# 2	0.217	16:02
Breath Volume		2.234	16:02
Air Blank		0.000	16:03
Control Test		0.077	16:03
Air Blank		0.000	16:04
Diagnostics Check		OK	16:04

Data Download Date/Time: 11/27/2018 10:08

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURPHY,JAVON,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MURPHY,JAVON
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 05:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:54

SUBJECT'S LAST NAME: PITTMAN

DATE/TIME OF TEST: 11/25/2018 06:38:49

Diagnostics Check		OK	06:32
Air Blank		0.000	06:33
Control Test		0.079	06:33
Air Blank		0.000	06:33
Subject Sample	# 1	0.104	06:34
Breath Volume		3.250	06:34
Air Blank		0.000	06:34
Air Blank		0.000	06:36
Subject Sample	# 2	0.100	06:37
Breath Volume		3.347	06:37
Air Blank		0.000	06:37
Control Test		0.080	06:38
Air Blank		0.000	06:38
Diagnostics Check		OK	06:38

Data Download Date/Time: 11/27/2018 10:08

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUNT,CHARLES,R
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: HUNT,CHARLES
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 10/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:46

SUBJECT'S LAST NAME: WEBB

DATE/TIME OF TEST: 11/26/2018 20:14:54

Diagnostics Check		OK	20:07
Air Blank		0.000	20:07
Control Test		0.080	20:08
Air Blank		0.000	20:08
Subject Sample	# 1	0.000	20:09
Breath Volume		1.113	20:09
Air Blank		0.000	20:09
Air Blank		0.000	20:11
Subject Sample	# 2	0.000	20:13
Breath Volume		1.289	20:13
Air Blank		0.000	20:13
Control Test		0.080	20:14
Air Blank		0.000	20:14
Diagnostics Check		OK	20:14

Data Download Date/Time: 11/27/2018 10:08

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FALGOUT,DAVID,G
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: BUTLER,G
ARRESTING OFFICER AGENCY: Live Oak PD
LAST AGENCY INSPECTION: 10/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 11/04/2018 23:18:20

Diagnostics Check		OK	23:12
Air Blank		0.000	23:12
Control Test		0.079	23:13
Air Blank		0.000	23:13
Subject Sample	# 1	0.200	23:13
Breath Volume		1.488	23:13
Air Blank		0.000	23:14
Air Blank		0.000	23:16
Subject Sample	# 2	0.192	23:16
Breath Volume		1.464	23:16
Air Blank		0.000	23:17
Control Test		0.077	23:17
Air Blank		0.000	23:18
Diagnostics Check		OK	23:18

Data Download Date/Time: 11/19/2018 14:01

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIEGEL,TIMOTHY,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: RIEGEL,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/08/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:13

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 11/11/2018 19:43:21

Diagnostics Check		OK	19:33	
Air Blank		0.000	19:34	
Control Test		0.079	19:34	
Air Blank		0.000	19:34	
Subject Sample	# 1	0.231	19:36	
Breath Volume		1.859	19:36	
Air Blank		0.000	19:37	
Air Blank		0.000	19:38	
Subject Sample	# 2	0.220	19:41	Volume Not Met
Breath Volume		0.886	19:41	
Air Blank			19:42	Purge Fail
Air Blank		0.000	19:43	

Data Download Date/Time: 11/19/2018 14:01

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FALGOUT,DAVID,G
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: MIARA,J
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 11/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: GREGORY

DATE/TIME OF TEST: 11/30/2018 21:11:48

Diagnostics Check		OK	21:05
Air Blank		0.000	21:05
Control Test		0.078	21:06
Air Blank		0.000	21:06
Subject Sample	# 1	0.134	21:07
Breath Volume		1.761	21:07
Air Blank		0.000	21:07
Air Blank		0.000	21:09
Subject Sample	# 2	0.133	21:10
Breath Volume		1.972	21:10
Air Blank		0.000	21:10
Control Test		0.078	21:11
Air Blank		0.000	21:11
Diagnostics Check		OK	21:11

Data Download Date/Time: 12/31/2018 11:55

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALEXANDER,LLOYD,A
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ALEXANDER,LLOYD
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG715202
DRY GAS EXPIRY DATE: 06/01/2019
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 11/01/2018 04:24:16

Diagnostics Check		OK	04:17
Air Blank		0.000	04:17
Control Test		0.082	04:18
Air Blank		0.000	04:18
Subject Sample	# 1	0.211	04:19
Breath Volume		1.597	04:19
Air Blank		0.000	04:20
Air Blank		0.000	04:21
Subject Sample	# 2	0.216	04:22
Breath Volume		1.812	04:22
Air Blank		0.000	04:23
Control Test		0.081	04:23
Air Blank		0.000	04:24
Diagnostics Check		OK	04:24

Data Download Date/Time: 11/03/2018 00:34

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARVOTH,DANIEL,P
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: HARVOTH,DANIEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG715202
DRY GAS EXPIRY DATE: 06/01/2019
VIOLATION CODE: DUI
ARREST TIME: 22:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: LAICHE

DATE/TIME OF TEST: 11/02/2018 22:48:19

Diagnostics Check		OK	22:41
Air Blank		0.000	22:41
Control Test		0.083	22:41
Air Blank		0.000	22:42
Subject Sample	# 1	0.186	22:43
Breath Volume		2.593	22:43
Air Blank		0.000	22:44
Air Blank		0.000	22:46
Subject Sample	# 2	0.187	22:46
Breath Volume		2.542	22:46
Air Blank		0.000	22:47
Control Test		0.082	22:47
Air Blank		0.000	22:48
Diagnostics Check		OK	22:48

Data Download Date/Time: 11/03/2018 00:34

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALEXANDER,LLOYD,A
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ALEXANDER,LLOYD
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG715202
DRY GAS EXPIRY DATE: 06/01/2019
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: MESSER

DATE/TIME OF TEST: 11/03/2018 04:01:35

Diagnostics Check		OK	03:49	
Air Blank		0.000	03:49	
Control Test		0.082	03:50	
Air Blank		0.000	03:50	
Subject Sample	# 1	0.195	03:53	Volume Not Met
Breath Volume		0.593	03:53	
Air Blank		0.000	03:54	
Air Blank		0.000	03:56	
Subject Sample	# 2	0.191	03:56	
Breath Volume		3.312	03:56	
Air Blank		0.000	03:57	
Air Blank		0.000	03:59	
Subject Sample	# 3	0.183	04:00	
Breath Volume		1.996	04:00	
Air Blank		0.000	04:00	
Control Test		0.081	04:00	
Air Blank		0.000	04:01	
Diagnostics Check		OK	04:01	

Data Download Date/Time: 11/30/2018 22:05

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,DANIEL,K
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: DAVIS,DANIEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG715202
DRY GAS EXPIRY DATE: 06/01/2019
VIOLATION CODE: DUI
ARREST TIME: 08:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:58

SUBJECT'S LAST NAME: BOURGEOIS

DATE/TIME OF TEST: 11/03/2018 09:28:21

Diagnostics Check		OK	09:22
Air Blank		0.000	09:22
Control Test		0.082	09:23
Air Blank		0.000	09:23
Subject Sample	# 1	0.000	09:24
Breath Volume		1.878	09:24
Air Blank		0.000	09:24
Air Blank		0.000	09:26
Subject Sample	# 2	0.000	09:26
Breath Volume		2.992	09:26
Air Blank		0.000	09:27
Control Test		0.082	09:27
Air Blank		0.000	09:28
Diagnostics Check		OK	09:28

Data Download Date/Time: 11/30/2018 22:05

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIRCH,JESSIE,E
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: BIRCH,JESSIE
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG715202
DRY GAS EXPIRY DATE: 06/01/2019
VIOLATION CODE: DUI
ARREST TIME: 11:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:08

SUBJECT'S LAST NAME: LUCIA

DATE/TIME OF TEST: 11/03/2018 12:35:25

Diagnostics Check		OK	12:29
Air Blank		0.000	12:29
Control Test		0.082	12:29
Air Blank		0.000	12:30
Subject Sample	# 1	0.177	12:30
Breath Volume		3.703	12:30
Air Blank		0.000	12:31
Air Blank		0.000	12:33
Subject Sample	# 2	0.180	12:33
Breath Volume		4.066	12:33
Air Blank		0.000	12:34
Control Test		0.082	12:34
Air Blank		0.000	12:35
Diagnostics Check		OK	12:35

Data Download Date/Time: 11/30/2018 22:05

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEINER,ZACHARY,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: STEINER,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG715202
DRY GAS EXPIRY DATE: 06/01/2019
VIOLATION CODE: DUI
ARREST TIME: 18:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: BARDIN

DATE/TIME OF TEST: 11/09/2018 19:27:29

Diagnostics Check		OK	19:24	
Air Blank		0.000	19:25	
Control Test		0.082	19:25	
Air Blank		0.000	19:25	
Subject Sample	# 1		19:26	Subject Test Refused
Breath Volume		0.000	19:26	
Air Blank		0.000	19:26	
Control Test		0.081	19:26	
Air Blank		0.000	19:27	
Diagnostics Check		OK	19:27	

Data Download Date/Time: 11/30/2018 22:05

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEINER,ZACHARY,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: STEINER,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG715202
DRY GAS EXPIRY DATE: 06/01/2019
VIOLATION CODE: DUI
ARREST TIME: 15:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:19

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 11/10/2018 16:46:05

Diagnostics Check		OK	16:39
Air Blank		0.000	16:40
Control Test		0.083	16:40
Air Blank		0.000	16:41
Subject Sample	# 1	0.000	16:41
Breath Volume		3.738	16:41
Air Blank		0.000	16:42
Air Blank		0.000	16:44
Subject Sample	# 2	0.000	16:44
Breath Volume		3.652	16:44
Air Blank		0.000	16:45
Control Test		0.083	16:45
Air Blank		0.000	16:46
Diagnostics Check		OK	16:46

Data Download Date/Time: 11/30/2018 22:05

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIVINGSTON,RON,C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LIVINGSTON,RON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG715202
DRY GAS EXPIRY DATE: 06/01/2019
VIOLATION CODE: DUI
ARREST TIME: 10:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:43

SUBJECT'S LAST NAME: BECTON

DATE/TIME OF TEST: 11/13/2018 11:10:28

Diagnostics Check		OK	11:04
Air Blank		0.000	11:04
Control Test		0.082	11:04
Air Blank		0.000	11:05
Subject Sample	# 1	0.000	11:06
Breath Volume		3.628	11:06
Air Blank		0.000	11:06
Air Blank		0.000	11:08
Subject Sample	# 2	0.000	11:09
Breath Volume		2.914	11:09
Air Blank		0.000	11:09
Control Test		0.082	11:09
Air Blank		0.000	11:10
Diagnostics Check		OK	11:10

Data Download Date/Time: 11/30/2018 22:05

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANEY,BRAD,A
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: BANEY,BRAD
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG715202
DRY GAS EXPIRY DATE: 06/01/2019
VIOLATION CODE: DUI
ARREST TIME: 21:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:28

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 11/14/2018 21:55:22

Diagnostics Check		OK	21:48
Air Blank		0.000	21:48
Control Test		0.082	21:48
Air Blank		0.000	21:49
Subject Sample	# 1	0.000	21:49
Breath Volume		1.562	21:49
Air Blank		0.000	21:50
Air Blank		0.000	21:52
Subject Sample	# 2	0.000	21:54
Breath Volume		2.406	21:54
Air Blank		0.000	21:54
Control Test		0.083	21:54
Air Blank		0.000	21:55
Diagnostics Check		OK	21:55

Data Download Date/Time: 11/30/2018 22:05

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER,RAYMOND,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SPENCER,RAYMOND
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG715202
DRY GAS EXPIRY DATE: 06/01/2019
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: JARRELL

DATE/TIME OF TEST: 11/16/2018 23:37:52

Diagnostics Check		OK	23:34	
Air Blank		0.000	23:34	
Control Test		0.082	23:35	
Air Blank		0.000	23:35	
Subject Sample	# 1		23:36	RFI Detect
Breath Volume		0.000	23:36	
Air Blank		0.000	23:36	
Control Test		0.083	23:37	
Air Blank		0.000	23:37	
Diagnostics Check		OK	23:37	

Data Download Date/Time: 11/30/2018 22:05

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER,RAYMOND,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SPENCER,RAYMOND
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG715202
DRY GAS EXPIRY DATE: 06/01/2019
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: JARRELL

DATE/TIME OF TEST: 11/16/2018 23:42:38

Diagnostics Check		OK	23:39	
Air Blank		0.000	23:39	
Control Test		0.083	23:39	
Air Blank		0.000	23:40	
Subject Sample	# 1		23:41	Subject Test Refused
Breath Volume		0.000	23:41	
Air Blank		0.000	23:41	
Control Test		0.082	23:42	
Air Blank		0.000	23:42	
Diagnostics Check		OK	23:42	

Data Download Date/Time: 11/30/2018 22:05

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KESSINGER,BRANDT,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: KESSINGER,B
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG715202
DRY GAS EXPIRY DATE: 06/01/2019
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 11/17/2018 03:55:57

Diagnostics Check		OK	03:44	
Air Blank		0.000	03:45	
Control Test		0.082	03:45	
Air Blank		0.000	03:45	
Subject Sample	# 1		03:48	No Sample Provided
Breath Volume		0.000	03:48	
Air Blank		0.000	03:49	
Air Blank		0.000	03:51	
Subject Sample	# 2		03:54	No Sample Provided
Breath Volume		0.000	03:54	
Air Blank		0.000	03:55	
Control Test		0.083	03:55	
Air Blank		0.000	03:55	
Diagnostics Check		OK	03:55	

Data Download Date/Time: 11/30/2018 22:05

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,COREY,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: JONES,COREY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG715202
DRY GAS EXPIRY DATE: 06/01/2019
VIOLATION CODE: DUI
ARREST TIME: 19:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:39

SUBJECT'S LAST NAME: ORTH

DATE/TIME OF TEST: 11/17/2018 21:07:04

Diagnostics Check		OK	21:00
Air Blank		0.000	21:00
Control Test		0.083	21:01
Air Blank		0.000	21:01
Subject Sample	# 1	0.000	21:02
Breath Volume		1.257	21:02
Air Blank		0.000	21:02
Air Blank		0.000	21:04
Subject Sample	# 2	0.000	21:05
Breath Volume		2.164	21:05
Air Blank		0.000	21:06
Control Test		0.082	21:06
Air Blank		0.000	21:06
Diagnostics Check		OK	21:07

Data Download Date/Time: 11/30/2018 22:05

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TARDIFF,FRANCIS,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: FUSZNER,THOMAS
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG715202
DRY GAS EXPIRY DATE: 06/01/2019
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: TARDIFF

DATE/TIME OF TEST: 11/18/2018 05:01:50

Diagnostics Check		OK	04:55
Air Blank		0.000	04:55
Control Test		0.082	04:55
Air Blank		0.000	04:56
Subject Sample	# 1	0.182	04:57
Breath Volume		3.921	04:57
Air Blank		0.000	04:57
Air Blank		0.000	04:59
Subject Sample	# 2	0.188	05:00
Breath Volume		3.671	05:00
Air Blank		0.000	05:00
Control Test		0.081	05:01
Air Blank		0.000	05:01
Diagnostics Check		OK	05:01

Data Download Date/Time: 11/30/2018 22:05

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALABRO,ALFRED,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CALABRO,ALFRED
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG715202
DRY GAS EXPIRY DATE: 06/01/2019
VIOLATION CODE: DUI
ARREST TIME: 16:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: FALLETTA

DATE/TIME OF TEST: 11/18/2018 17:34:10

Diagnostics Check		OK	17:21	
Air Blank		0.000	17:21	
Control Test		0.081	17:22	
Air Blank		0.000	17:22	
Subject Sample	# 1	0.239	17:25	Volume Not Met
Breath Volume		0.171	17:25	
Air Blank		0.000	17:26	
Air Blank		0.000	17:28	
Subject Sample	# 2	0.217	17:29	
Breath Volume		3.031	17:29	
Air Blank		0.000	17:30	
Air Blank		0.000	17:32	
Subject Sample	# 3	0.224	17:32	
Breath Volume		2.421	17:32	
Air Blank		0.000	17:33	
Control Test		0.080	17:33	
Air Blank		0.000	17:34	
Diagnostics Check		OK	17:34	

Data Download Date/Time: 11/30/2018 22:05

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,STEPHEN,W
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: QUICK,STEPHEN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG715202
DRY GAS EXPIRY DATE: 06/01/2019
VIOLATION CODE: DUI
ARREST TIME: 05:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:28

SUBJECT'S LAST NAME: OWENS

DATE/TIME OF TEST: 11/20/2018 06:56:29

Diagnostics Check		OK	06:49
Air Blank		0.000	06:49
Control Test		0.082	06:49
Air Blank		0.000	06:50
Subject Sample	# 1	0.199	06:51
Breath Volume		1.910	06:51
Air Blank		0.000	06:52
Air Blank		0.000	06:54
Subject Sample	# 2	0.196	06:54
Breath Volume		1.464	06:54
Air Blank		0.000	06:55
Control Test		0.081	06:55
Air Blank		0.000	06:56
Diagnostics Check		OK	06:56

Data Download Date/Time: 11/30/2018 22:05

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWTON,CARSON,H
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: HOWTON,CARSON
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG715202
DRY GAS EXPIRY DATE: 06/01/2019
VIOLATION CODE: DUI
ARREST TIME: 16:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:37

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 11/21/2018 17:04:33

Diagnostics Check		OK	16:57
Air Blank		0.000	16:58
Control Test		0.083	16:58
Air Blank		0.000	16:59
Subject Sample	# 1	0.000	17:00
Breath Volume		2.718	17:00
Air Blank		0.000	17:00
Air Blank		0.000	17:02
Subject Sample	# 2	0.000	17:03
Breath Volume		2.511	17:03
Air Blank		0.000	17:03
Control Test		0.083	17:03
Air Blank		0.000	17:04
Diagnostics Check		OK	17:04

Data Download Date/Time: 11/30/2018 22:05

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLEMENTS,WILLIAM,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CLEMENTS,WILLIAM
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG715202
DRY GAS EXPIRY DATE: 06/01/2019
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 11/21/2018 23:40:59

Diagnostics Check		OK	23:37	
Air Blank		0.000	23:37	
Control Test		0.082	23:37	
Air Blank		0.000	23:38	
Subject Sample	# 1		23:39	Subject Test Refused
Breath Volume		0.000	23:39	
Air Blank		0.000	23:40	
Control Test		0.082	23:40	
Air Blank		0.000	23:40	
Diagnostics Check		OK	23:40	

Data Download Date/Time: 11/30/2018 22:05

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERSON,GABRIEL,A
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ROBERSON,GABRIEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG715202
DRY GAS EXPIRY DATE: 06/01/2019
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 11/24/2018 23:56:29

Diagnostics Check		OK	23:50	
Air Blank		0.000	23:51	
Control Test		0.082	23:51	
Air Blank		0.000	23:51	
Subject Sample	# 1		23:54	Subject Test Refused
Breath Volume		0.121	23:54	
Air Blank		0.000	23:55	
Control Test		0.081	23:55	
Air Blank		0.000	23:56	
Diagnostics Check		OK	23:56	

Data Download Date/Time: 11/30/2018 22:05

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERSON,GABRIEL,A
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ROBERSON,GABRIEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG715202
DRY GAS EXPIRY DATE: 06/01/2019
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 11/25/2018 00:19:11

Diagnostics Check		OK	00:15	
Air Blank		0.000	00:15	
Control Test		0.080	00:15	
Air Blank		0.000	00:16	
Subject Sample	# 1		00:17	Subject Test Refused
Breath Volume		0.433	00:17	
Air Blank		0.000	00:18	
Control Test		0.081	00:18	
Air Blank		0.000	00:19	
Diagnostics Check		OK	00:19	

Data Download Date/Time: 11/30/2018 22:05

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COFELT,DYLAN,H
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: COFFELT,DYLAN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG715202
DRY GAS EXPIRY DATE: 06/01/2019
VIOLATION CODE: DUI
ARREST TIME: 17:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:07

SUBJECT'S LAST NAME: HASELHORST

DATE/TIME OF TEST: 11/26/2018 18:36:15

Diagnostics Check		OK	18:27
Air Blank		0.000	18:27
Control Test		0.083	18:28
Air Blank		0.000	18:28
Subject Sample	# 1	0.335	18:28
Breath Volume		2.832	18:28
Air Blank		0.000	18:29
Air Blank		0.000	18:31
Subject Sample	# 2	0.313	18:31
Breath Volume		1.632	18:31
Air Blank		0.000	18:32
Air Blank		0.000	18:34
Subject Sample	# 3	0.313	18:34
Breath Volume		1.500	18:34
Air Blank		0.000	18:35
Control Test		0.082	18:35
Air Blank		0.000	18:36
Diagnostics Check		OK	18:36

Data Download Date/Time: 11/30/2018 22:05

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLEMENTS,WILLIAM,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CLEMENTS,WILLIAM
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG715202
DRY GAS EXPIRY DATE: 06/01/2019
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: FEERICK

DATE/TIME OF TEST: 11/27/2018 02:19:19

Diagnostics Check		OK	02:15	
Air Blank		0.000	02:15	
Control Test		0.083	02:16	
Air Blank		0.000	02:16	
Subject Sample	# 1		02:17	Subject Test Refused
Breath Volume		0.000	02:17	
Air Blank		0.000	02:18	
Control Test		0.082	02:18	
Air Blank		0.000	02:19	
Diagnostics Check		OK	02:19	

Data Download Date/Time: 11/30/2018 22:05

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMIREZ,MICHAEL,E
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: RAMIREZ,MICHAEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG715202
DRY GAS EXPIRY DATE: 06/01/2019
VIOLATION CODE: DUI
ARREST TIME: 21:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: LONGTINE

DATE/TIME OF TEST: 11/29/2018 22:34:55

Diagnostics Check		OK	22:23	
Air Blank		0.000	22:23	
Control Test		0.081	22:24	
Air Blank		0.000	22:24	
Subject Sample	# 1		22:27	No Sample Provided
Breath Volume		0.000	22:27	
Air Blank		0.000	22:28	
Air Blank		0.000	22:30	
Subject Sample	# 2		22:33	No Sample Provided
Breath Volume		0.000	22:33	
Air Blank		0.000	22:34	
Control Test		0.082	22:34	
Air Blank		0.000	22:34	
Diagnostics Check		OK	22:34	

Data Download Date/Time: 11/30/2018 22:05

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARGITT,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: VOLNER,THOMAS
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG715202
DRY GAS EXPIRY DATE: 06/01/2019
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: PAGE

DATE/TIME OF TEST: 11/29/2018 23:31:32

Diagnostics Check		OK	23:26	
Air Blank		0.000	23:26	
Control Test		0.081	23:27	
Air Blank		0.000	23:27	
Subject Sample	# 1		23:30	Subject Test Refused
Breath Volume		0.000	23:30	
Air Blank		0.000	23:30	
Control Test		0.082	23:30	
Air Blank		0.000	23:31	
Diagnostics Check		OK	23:31	

Data Download Date/Time: 11/30/2018 22:05

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIDERGREN,DAVID,A
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: BORCHARDT,GARY
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 10/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 13:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:35

SUBJECT'S LAST NAME: ESTEP

DATE/TIME OF TEST: 11/05/2018 14:44:57

Diagnostics Check		OK	14:38
Air Blank		0.000	14:38
Control Test		0.080	14:38
Air Blank		0.000	14:39
Subject Sample	# 1	0.271	14:40
Breath Volume		2.218	14:40
Air Blank		0.000	14:40
Air Blank		0.000	14:42
Subject Sample	# 2	0.262	14:43
Breath Volume		1.738	14:43
Air Blank		0.000	14:44
Control Test		0.079	14:44
Air Blank		0.000	14:44
Diagnostics Check		OK	14:44

Data Download Date/Time: 01/11/2019 11:56

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATES,BRIAN,C
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BATES,BRIAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: LIVINGSTON

DATE/TIME OF TEST: 11/08/2018 23:09:27

Diagnostics Check		OK	23:03
Air Blank		0.000	23:03
Control Test		0.082	23:04
Air Blank		0.000	23:04
Subject Sample	# 1	0.217	23:04
Breath Volume		2.156	23:04
Air Blank		0.000	23:05
Air Blank		0.000	23:07
Subject Sample	# 2	0.208	23:07
Breath Volume		1.867	23:07
Air Blank		0.000	23:08
Control Test		0.078	23:08
Air Blank		0.000	23:09
Diagnostics Check		OK	23:09

Data Download Date/Time: 01/11/2019 11:56

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CREWS,MICHAEL,B
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: MONTEMURNO,MATT
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 10/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 11/11/2018 01:49:07

Diagnostics Check		OK	01:42
Air Blank		0.000	01:42
Control Test		0.081	01:43
Air Blank		0.000	01:43
Subject Sample	# 1	0.217	01:44
Breath Volume		2.058	01:44
Air Blank		0.000	01:45
Air Blank		0.000	01:46
Subject Sample	# 2	0.230	01:47
Breath Volume		2.929	01:47
Air Blank		0.000	01:48
Control Test		0.080	01:48
Air Blank		0.000	01:49
Diagnostics Check		OK	01:49

Data Download Date/Time: 01/11/2019 11:56

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARVEY,DAVID,A
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: HARVEY,DAVID
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 10/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 11/11/2018 03:09:41

Diagnostics Check		OK	03:01	
Air Blank		0.000	03:01	
Control Test		0.081	03:01	
Air Blank		0.000	03:02	
Subject Sample	# 1		03:05	No Sample Provided
Breath Volume		0.000	03:05	
Air Blank		0.000	03:05	
Air Blank		0.000	03:07	
Subject Sample	# 2		03:08	Subject Test Refused
Breath Volume		0.000	03:08	
Air Blank		0.000	03:08	
Control Test		0.081	03:09	
Air Blank		0.000	03:09	
Diagnostics Check		OK	03:09	

Data Download Date/Time: 01/11/2019 11:56

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARVEY,DAVID,A
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: MONTEMURNO,MATTHEW
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 10/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 04:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 11/11/2018 04:56:44

Diagnostics Check		OK	04:48	
Air Blank		0.000	04:48	
Control Test		0.081	04:48	
Air Blank		0.000	04:49	
Subject Sample	# 1		04:52	No Sample Provided
Breath Volume		0.000	04:52	
Air Blank		0.000	04:53	
Air Blank		0.000	04:55	
Subject Sample	# 2		04:55	Subject Test Refused
Breath Volume		0.000	04:55	
Air Blank		0.000	04:55	
Control Test		0.081	04:56	
Air Blank		0.000	04:56	
Diagnostics Check		OK	04:56	

Data Download Date/Time: 01/11/2019 11:56

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHELAN,MATTHEW,S
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: PHELAN,MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 11:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:20

SUBJECT'S LAST NAME: SKARZYNSKI

DATE/TIME OF TEST: 11/12/2018 12:46:18

Diagnostics Check		OK	12:40
Air Blank		0.000	12:40
Control Test		0.081	12:41
Air Blank		0.000	12:41
Subject Sample	# 1	0.000	12:42
Breath Volume		2.328	12:42
Air Blank		0.000	12:42
Air Blank		0.000	12:44
Subject Sample	# 2	0.000	12:44
Breath Volume		2.363	12:44
Air Blank		0.000	12:45
Control Test		0.080	12:45
Air Blank		0.000	12:46
Diagnostics Check		OK	12:46

Data Download Date/Time: 01/11/2019 11:56

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAURES,BRANDYN,K
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BATES,BRIAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/30/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: ROWELL

DATE/TIME OF TEST: 11/13/2018 01:53:07

Diagnostics Check		OK	01:46
Air Blank		0.000	01:46
Control Test		0.081	01:46
Air Blank		0.000	01:47
Subject Sample	# 1	0.142	01:48
Breath Volume		2.207	01:48
Air Blank		0.000	01:49
Air Blank		0.000	01:50
Subject Sample	# 2	0.148	01:51
Breath Volume		1.644	01:51
Air Blank		0.000	01:52
Control Test		0.081	01:52
Air Blank		0.000	01:53
Diagnostics Check		OK	01:53

Data Download Date/Time: 01/11/2019 11:56

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHELAN,MATTHEW,S
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MCFARLAND,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/30/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 05:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:40

SUBJECT'S LAST NAME: KISSNER

DATE/TIME OF TEST: 11/13/2018 07:06:53

Diagnostics Check		OK	07:00
Air Blank		0.000	07:01
Control Test		0.081	07:01
Air Blank		0.000	07:01
Subject Sample	# 1	0.140	07:02
Breath Volume		3.191	07:02
Air Blank		0.000	07:03
Air Blank		0.000	07:04
Subject Sample	# 2	0.137	07:05
Breath Volume		3.183	07:05
Air Blank		0.000	07:05
Control Test		0.080	07:06
Air Blank		0.000	07:06
Diagnostics Check		OK	07:06

Data Download Date/Time: 01/11/2019 11:56

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHELAN,MATTHEW,S
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: PHELAN,MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 11/17/2018 04:11:36

Diagnostics Check		OK	04:05
Air Blank		0.000	04:05
Control Test		0.082	04:06
Air Blank		0.000	04:06
Subject Sample	# 1	0.185	04:07
Breath Volume		2.804	04:07
Air Blank		0.000	04:07
Air Blank		0.000	04:09
Subject Sample	# 2	0.184	04:10
Breath Volume		3.027	04:10
Air Blank		0.000	04:10
Control Test		0.079	04:11
Air Blank		0.000	04:11
Diagnostics Check		OK	04:11

Data Download Date/Time: 01/11/2019 11:56

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINSON,STUART,P
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: ROBINSON,STUART
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 10/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: ODEN

DATE/TIME OF TEST: 11/17/2018 04:21:38

Diagnostics Check		OK	04:14
Air Blank		0.000	04:15
Control Test		0.080	04:15
Air Blank		0.000	04:16
Subject Sample	# 1	0.176	04:16
Breath Volume		2.589	04:16
Air Blank		0.000	04:17
Air Blank		0.000	04:19
Subject Sample	# 2	0.168	04:19
Breath Volume		1.894	04:19
Air Blank		0.000	04:20
Control Test		0.079	04:21
Air Blank		0.000	04:21
Diagnostics Check		OK	04:21

Data Download Date/Time: 01/11/2019 11:56

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,DANIEL,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MARTIN,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 05:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:58

SUBJECT'S LAST NAME: BONDS

DATE/TIME OF TEST: 11/17/2018 06:21:47

Diagnostics Check		OK	06:18	
Air Blank		0.000	06:19	
Control Test		0.082	06:19	
Air Blank		0.000	06:20	
Subject Sample	# 1		06:20	Subject Test Refused
Breath Volume		0.000	06:20	
Air Blank		0.000	06:20	
Control Test		0.082	06:21	
Air Blank		0.000	06:21	
Diagnostics Check		OK	06:21	

Data Download Date/Time: 01/11/2019 11:56

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATES,BRIAN,C
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: LARA,ALEX
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 18:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: VADNAIS

DATE/TIME OF TEST: 11/21/2018 20:08:03

Diagnostics Check		OK	20:01
Air Blank		0.000	20:01
Control Test		0.081	20:01
Air Blank		0.000	20:02
Subject Sample	# 1	0.275	20:03
Breath Volume		2.187	20:03
Air Blank		0.000	20:04
Air Blank		0.000	20:06
Subject Sample	# 2	0.256	20:06
Breath Volume		1.863	20:06
Air Blank		0.000	20:07
Control Test		0.077	20:07
Air Blank		0.000	20:07
Diagnostics Check		OK	20:08

Data Download Date/Time: 01/11/2019 11:56

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAURES,BRANDYN,K
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BATES,BRIAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 11/22/2018 03:00:39

Diagnostics Check		OK	02:54
Air Blank		0.000	02:54
Control Test		0.080	02:55
Air Blank		0.000	02:55
Subject Sample	# 1	0.000	02:56
Breath Volume		2.050	02:56
Air Blank		0.000	02:56
Air Blank		0.000	02:58
Subject Sample	# 2	0.000	02:59
Breath Volume		1.472	02:59
Air Blank		0.000	02:59
Control Test		0.081	03:00
Air Blank		0.000	03:00
Diagnostics Check		OK	03:00

Data Download Date/Time: 01/11/2019 11:56

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SINGLETARY,MATHEW,C
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SINGLETARY,MATHEW
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: MUNG

DATE/TIME OF TEST: 11/25/2018 01:27:17

Diagnostics Check		OK	01:19
Air Blank		0.000	01:19
Control Test		0.081	01:19
Air Blank		0.000	01:20
Subject Sample	# 1	0.169	01:22
Breath Volume		1.394	01:22
Air Blank		0.000	01:23
Air Blank		0.000	01:24
Subject Sample	# 2	0.166	01:25
Breath Volume		1.210	01:25
Air Blank		0.000	01:26
Control Test		0.080	01:26
Air Blank		0.000	01:27
Diagnostics Check		OK	01:27

Data Download Date/Time: 01/11/2019 11:56

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000776

INSTRUMENT REGISTERED TO: LAFAYETTE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIS,CLEVELAND,D
BREATH TEST OPERATOR AGENCY: Lafayette County SO
ARRESTING OFFICER: WILLIS,CLEVELAND
ARRESTING OFFICER AGENCY: Lafayette County SO
LAST AGENCY INSPECTION: 10/24/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1761808A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: ELIASON

DATE/TIME OF TEST: 11/06/2018 18:14:25

Diagnostics Check		OK	18:11	
Air Blank		0.000	18:11	
Control Test		0.081	18:11	
Air Blank		0.000	18:12	
Subject Sample	# 1		18:12	Slope Not Met
Breath Volume		1.187	18:12	
Air Blank		0.000	18:13	
Control Test		0.081	18:13	
Air Blank		0.000	18:14	
Diagnostics Check		OK	18:14	

Data Download Date/Time: 12/11/2018 13:12

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000776

INSTRUMENT REGISTERED TO: LAFAYETTE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIS,CLEVELAND,D
BREATH TEST OPERATOR AGENCY: Lafayette County SO
ARRESTING OFFICER: WILLIS,CLEVELAND
ARRESTING OFFICER AGENCY: Lafayette County SO
LAST AGENCY INSPECTION: 10/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1761808A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: ELIASON

DATE/TIME OF TEST: 11/06/2018 18:24:27

Diagnostics Check		OK	18:17
Air Blank		0.000	18:17
Control Test		0.080	18:17
Air Blank		0.000	18:18
Subject Sample	# 1	0.341	18:19
Breath Volume		1.687	18:19
Air Blank		0.000	18:19
Air Blank		0.000	18:21
Subject Sample	# 2	0.335	18:22
Breath Volume		2.308	18:22
Air Blank		0.000	18:23
Control Test		0.079	18:23
Air Blank		0.000	18:24
Diagnostics Check		OK	18:24

Data Download Date/Time: 12/11/2018 13:12

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000779

INSTRUMENT REGISTERED TO: MADISON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSCHETTO,ANTHONY,C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: JOOST,MARK
ARRESTING OFFICER AGENCY: Madison PD
LAST AGENCY INSPECTION: 10/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A3
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 12:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:00

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 11/10/2018 13:37:35

Diagnostics Check		OK	13:30
Air Blank		0.000	13:31
Control Test		0.080	13:31
Air Blank		0.000	13:32
Subject Sample	# 1	0.285	13:32
Breath Volume		2.476	13:32
Air Blank		0.000	13:33
Air Blank		0.000	13:35
Subject Sample	# 2	0.274	13:35
Breath Volume		1.433	13:35
Air Blank		0.000	13:36
Control Test		0.080	13:36
Air Blank		0.000	13:37
Diagnostics Check		OK	13:37

Data Download Date/Time: 11/14/2018 10:43

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NATIVIDAD,DANIEL,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: NATIVIDAD,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 11/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:09

SUBJECT'S LAST NAME: HOWELL

DATE/TIME OF TEST: 11/15/2018 19:31:18

Diagnostics Check		OK	19:29	
Air Blank		0.000	19:29	
Control Test		0.078	19:30	
Air Blank		0.000	19:30	
Subject Sample	# 1		19:31	RFI Detect
Breath Volume		0.000	19:31	
Air Blank			19:31	RFI Detect
Air Blank			19:31	RFI Detect

Data Download Date/Time: 01/07/2019 15:29

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NATIVIDAD,DANIEL,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: NATIVIDAD,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 11/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:09

SUBJECT'S LAST NAME: HOWELL

DATE/TIME OF TEST: 11/15/2018 19:39:24

Diagnostics Check		OK	19:32
Air Blank		0.000	19:33
Control Test		0.080	19:33
Air Blank		0.000	19:34
Subject Sample	# 1	0.000	19:34
Breath Volume		4.371	19:34
Air Blank		0.000	19:35
Air Blank		0.000	19:37
Subject Sample	# 2	0.000	19:38
Breath Volume		3.476	19:38
Air Blank		0.000	19:38
Control Test		0.079	19:38
Air Blank		0.000	19:39
Diagnostics Check		OK	19:39

Data Download Date/Time: 01/07/2019 15:29

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,MITCHELL,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MORRIS,DANIEL
ARRESTING OFFICER AGENCY: Holmes County SO
LAST AGENCY INSPECTION: 10/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 461951
DRY GAS EXPIRY DATE: 06/15/2019
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 11/04/2018 23:45:41

Diagnostics Check		OK	23:43	
Air Blank		0.000	23:43	
Control Test		0.081	23:44	
Air Blank		0.000	23:44	
Subject Sample	# 1		23:45	Sequence Aborted
Breath Volume		0.000	23:45	
Air Blank		0.000	23:45	

Data Download Date/Time: 11/05/2018 12:10

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEFF,STEPHEN,
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: BANKS,SEAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 11/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 17:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:35

SUBJECT'S LAST NAME: KERRIGAN

DATE/TIME OF TEST: 11/13/2018 18:00:08

Diagnostics Check		OK	17:57	
Air Blank		0.000	17:57	
Control Test		0.079	17:58	
Air Blank		0.000	17:58	
Subject Sample	# 1		17:58	Subject Test Refused
Breath Volume		0.000	17:58	
Air Blank		0.000	17:59	
Control Test		0.079	17:59	
Air Blank		0.000	18:00	
Diagnostics Check		OK	18:00	

Data Download Date/Time: 12/17/2018 14:56

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANKS,SEAN,C
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: BANKS,SEAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 11/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 15:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: KEENAN

DATE/TIME OF TEST: 11/18/2018 16:23:50

Diagnostics Check		OK	16:21	
Air Blank		0.000	16:21	
Control Test		0.080	16:21	
Air Blank		0.000	16:22	
Subject Sample	# 1		16:22	Subject Test Refused
Breath Volume		0.000	16:22	
Air Blank		0.000	16:22	
Control Test		0.080	16:23	
Air Blank		0.000	16:23	
Diagnostics Check		OK	16:23	

Data Download Date/Time: 12/17/2018 14:56

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEFF,STEPHEN,N
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: BANKS,SEAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 11/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 13:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:53

SUBJECT'S LAST NAME: ALONSO

DATE/TIME OF TEST: 11/21/2018 14:21:11

Diagnostics Check		OK	14:14
Air Blank		0.000	14:15
Control Test		0.079	14:15
Air Blank		0.000	14:16
Subject Sample	# 1	0.236	14:16
Breath Volume		2.605	14:16
Air Blank		0.000	14:17
Air Blank		0.000	14:19
Subject Sample	# 2	0.232	14:19
Breath Volume		3.207	14:19
Air Blank		0.000	14:20
Control Test		0.080	14:20
Air Blank		0.000	14:21
Diagnostics Check		OK	14:21

Data Download Date/Time: 12/17/2018 14:56

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,RICHARD,D
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: MONDRAGON,LIONEL
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: CURRIER

DATE/TIME OF TEST: 11/01/2018 03:34:48

Diagnostics Check		OK	03:28
Air Blank		0.000	03:28
Control Test		0.078	03:28
Air Blank		0.000	03:29
Subject Sample	# 1	0.171	03:30
Breath Volume		4.210	03:30
Air Blank		0.000	03:30
Air Blank		0.000	03:32
Subject Sample	# 2	0.161	03:33
Breath Volume		4.394	03:33
Air Blank		0.000	03:33
Control Test		0.078	03:34
Air Blank		0.000	03:34
Diagnostics Check		OK	03:34

Data Download Date/Time: 11/29/2018 10:41

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FENTON,WILLIAM,B
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: FENTON,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: EISENBREY

DATE/TIME OF TEST: 11/04/2018 22:12:32

Diagnostics Check		OK	22:04
Air Blank		0.000	22:05
Control Test		0.079	22:05
Air Blank		0.000	22:06
Subject Sample	# 1	0.268	22:08
Breath Volume		1.835	22:08
Air Blank		0.000	22:08
Air Blank		0.000	22:10
Subject Sample	# 2	0.266	22:10
Breath Volume		1.511	22:10
Air Blank		0.000	22:11
Control Test		0.078	22:11
Air Blank		0.000	22:12
Diagnostics Check		OK	22:12

Data Download Date/Time: 11/29/2018 10:41

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTON,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: ANTON,NICHOLAS
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: ROBLES-REYES

DATE/TIME OF TEST: 11/05/2018 04:12:00

Diagnostics Check		OK	04:05
Air Blank		0.000	04:05
Control Test		0.079	04:05
Air Blank		0.000	04:06
Subject Sample	# 1	0.206	04:07
Breath Volume		1.277	04:07
Air Blank		0.000	04:07
Air Blank		0.000	04:09
Subject Sample	# 2	0.200	04:10
Breath Volume		1.558	04:10
Air Blank		0.000	04:11
Control Test		0.078	04:11
Air Blank		0.000	04:11
Diagnostics Check		OK	04:12

Data Download Date/Time: 11/29/2018 10:41

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: DRAWDY,LAINE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: DIONNE

DATE/TIME OF TEST: 11/06/2018 01:43:45

Diagnostics Check		OK	01:37
Air Blank		0.000	01:37
Control Test		0.079	01:37
Air Blank		0.000	01:38
Subject Sample	# 1	0.283	01:39
Breath Volume		1.242	01:39
Air Blank		0.000	01:39
Air Blank		0.000	01:41
Subject Sample	# 2	0.284	01:42
Breath Volume		2.378	01:42
Air Blank		0.000	01:42
Control Test		0.078	01:43
Air Blank		0.000	01:43
Diagnostics Check		OK	01:43

Data Download Date/Time: 11/29/2018 10:41

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTON,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: ANTON,NICHOLAS
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:27

SUBJECT'S LAST NAME: STARBUCK

DATE/TIME OF TEST: 11/06/2018 21:02:15

Diagnostics Check		OK	20:58	
Air Blank		0.000	20:58	
Control Test		0.079	20:59	
Air Blank		0.000	20:59	
Subject Sample	# 1		21:00	RFI Detect
Breath Volume		0.945	21:00	
Air Blank		0.000	21:01	
Control Test		0.078	21:01	
Air Blank		0.000	21:02	
Diagnostics Check		OK	21:02	

Data Download Date/Time: 11/29/2018 10:41

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTON,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: ANTON,NICHOLAS
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:27

SUBJECT'S LAST NAME: STARBUCK

DATE/TIME OF TEST: 11/06/2018 21:14:20

Diagnostics Check		OK	21:04
Air Blank		0.000	21:04
Control Test		0.078	21:04
Air Blank		0.000	21:05
Subject Sample	# 1	0.179	21:08
Breath Volume		1.414	21:08
Air Blank		0.000	21:09
Air Blank		0.000	21:11
Subject Sample	# 2	0.171	21:12
Breath Volume		1.167	21:12
Air Blank		0.000	21:13
Control Test		0.077	21:13
Air Blank		0.000	21:14
Diagnostics Check		OK	21:14

Data Download Date/Time: 11/29/2018 10:41

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: DRAWDY,LAINE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: MCINTOSH

DATE/TIME OF TEST: 11/10/2018 00:45:44

Diagnostics Check		OK	00:39
Air Blank		0.000	00:39
Control Test		0.079	00:40
Air Blank		0.000	00:40
Subject Sample	# 1	0.194	00:41
Breath Volume		2.023	00:41
Air Blank		0.000	00:41
Air Blank		0.000	00:43
Subject Sample	# 2	0.193	00:44
Breath Volume		1.500	00:44
Air Blank		0.000	00:44
Control Test		0.078	00:45
Air Blank		0.000	00:45
Diagnostics Check		OK	00:45

Data Download Date/Time: 11/29/2018 10:41

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STOKES,SHAYNE,R
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: STOKS,SHAYNE
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:05

SUBJECT'S LAST NAME: CAUSSADE

DATE/TIME OF TEST: 11/10/2018 16:32:41

Diagnostics Check		OK	16:25
Air Blank		0.000	16:26
Control Test		0.079	16:26
Air Blank		0.000	16:26
Subject Sample	# 1	0.136	16:27
Breath Volume		1.507	16:27
Air Blank		0.000	16:28
Air Blank		0.000	16:30
Subject Sample	# 2	0.128	16:31
Breath Volume		1.160	16:31
Air Blank		0.000	16:31
Control Test		0.078	16:32
Air Blank		0.000	16:32
Diagnostics Check		OK	16:32

Data Download Date/Time: 11/29/2018 10:41

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: MONDRAGON,LIONEL
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: MENDEZ

DATE/TIME OF TEST: 11/11/2018 04:24:45

Diagnostics Check		OK	04:21	
Air Blank		0.000	04:22	
Control Test		0.078	04:22	
Air Blank		0.000	04:22	
Subject Sample	# 1		04:23	Subject Test Refused
Breath Volume		0.000	04:23	
Air Blank		0.000	04:23	
Control Test		0.079	04:24	
Air Blank		0.000	04:24	
Diagnostics Check		OK	04:24	

Data Download Date/Time: 11/29/2018 10:41

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEER,BRADLEY,J
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: MEER,BRADLEY
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:03

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 11/12/2018 21:41:39

Diagnostics Check		OK	21:38	
Air Blank		0.000	21:38	
Control Test		0.079	21:38	
Air Blank		0.000	21:39	
Subject Sample	# 1		21:40	RFI Detect
Breath Volume		0.000	21:40	
Air Blank		0.000	21:41	
Control Test		0.078	21:41	RFI Detect
Air Blank			21:41	RFI Detect

Data Download Date/Time: 11/29/2018 10:41

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GENTRY,JOSHUA,J
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: DOZIER,TERRELL
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: SINGLETON

DATE/TIME OF TEST: 11/13/2018 02:33:41

Diagnostics Check		OK	02:22	
Air Blank		0.000	02:22	
Control Test		0.079	02:23	
Air Blank		0.000	02:23	
Subject Sample	# 1		02:26	No Sample Provided
Breath Volume		0.000	02:26	
Air Blank		0.000	02:27	
Air Blank		0.000	02:29	
Subject Sample	# 2		02:32	No Sample Provided
Breath Volume		0.000	02:32	
Air Blank		0.000	02:32	
Control Test		0.079	02:33	
Air Blank		0.000	02:33	
Diagnostics Check		OK	02:33	

Data Download Date/Time: 11/29/2018 10:41

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTON,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: LIMB,HUNTER
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:08

SUBJECT'S LAST NAME: BOYD

DATE/TIME OF TEST: 11/13/2018 21:59:51

Diagnostics Check		OK	21:49
Air Blank		0.000	21:49
Control Test		0.079	21:50
Air Blank		0.000	21:50
Subject Sample	# 1	0.267	21:51
Breath Volume		1.195	21:51
Air Blank		0.000	21:51
Air Blank		0.000	21:53
Subject Sample	# 2	0.239	21:55
Breath Volume		1.574	21:55
Air Blank		0.000	21:55
Air Blank		0.000	21:57
Subject Sample	# 3	0.246	21:58
Breath Volume		2.398	21:58
Air Blank		0.000	21:58
Control Test		0.078	21:59
Air Blank		0.000	21:59
Diagnostics Check		OK	21:59

Data Download Date/Time: 11/29/2018 10:41

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,CODY,L
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: BROWN,CODY
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 10:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: HOLMES

DATE/TIME OF TEST: 11/17/2018 00:28:43

Diagnostics Check	OK	00:27	
Air Blank	0.000	00:27	
Control Test	0.078	00:28	
Air Blank		00:28	Improper Sample
Air Blank		00:28	Improper Sample

Data Download Date/Time: 11/29/2018 10:41

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,CODY,L
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: BROWN,CODY
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 10:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: HOLMES

DATE/TIME OF TEST: 11/17/2018 00:28:12

Diagnostics Check	OK	00:27	
Air Blank		00:27	Ambient Fail
Air Blank	0.000	00:28	

Data Download Date/Time: 11/29/2018 10:41

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,CODY,L
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: BROWN,CODY
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: HOLMES

DATE/TIME OF TEST: 11/17/2018 00:35:38

Diagnostics Check		OK	00:29
Air Blank		0.000	00:30
Control Test		0.078	00:30
Air Blank		0.000	00:30
Subject Sample	# 1	0.123	00:31
Breath Volume		1.722	00:31
Air Blank		0.000	00:31
Air Blank		0.000	00:33
Subject Sample	# 2	0.119	00:34
Breath Volume		1.757	00:34
Air Blank		0.000	00:34
Control Test		0.078	00:35
Air Blank		0.000	00:35
Diagnostics Check		OK	00:35

Data Download Date/Time: 11/29/2018 10:41

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALTENBURG,MARCUS,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: LIMB,HUNTER
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: ANTILL

DATE/TIME OF TEST: 11/18/2018 04:15:55

Diagnostics Check		OK	04:01	
Air Blank		0.000	04:02	
Control Test		0.079	04:02	
Air Blank		0.000	04:02	
Subject Sample	# 1	0.100	04:06	Volume Not Met
Breath Volume		0.171	04:06	
Air Blank		0.000	04:06	
Air Blank		0.000	04:08	
Subject Sample	# 2	0.129	04:10	
Breath Volume		1.117	04:10	
Air Blank		0.000	04:11	
Air Blank		0.000	04:12	
Subject Sample	# 3	0.130	04:14	
Breath Volume		1.460	04:14	
Air Blank		0.000	04:15	
Control Test		0.077	04:15	
Air Blank		0.000	04:15	
Diagnostics Check		OK	04:15	

Data Download Date/Time: 11/29/2018 10:41

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALTENBURG,MARCUS,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: WALTENBURG,MARCUS
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: BURNS

DATE/TIME OF TEST: 11/19/2018 02:49:40

Diagnostics Check		OK	02:43
Air Blank		0.000	02:43
Control Test		0.079	02:43
Air Blank		0.000	02:44
Subject Sample	# 1	0.202	02:45
Breath Volume		3.000	02:45
Air Blank		0.000	02:45
Air Blank		0.000	02:47
Subject Sample	# 2	0.202	02:48
Breath Volume		3.007	02:48
Air Blank		0.000	02:48
Control Test		0.078	02:49
Air Blank		0.000	02:49
Diagnostics Check		OK	02:49

Data Download Date/Time: 11/29/2018 10:41

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARNER,TODD,A
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: DENEAU,ALEX
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 13:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:44

SUBJECT'S LAST NAME: CULBERTSON

DATE/TIME OF TEST: 11/20/2018 16:11:12

Diagnostics Check		OK	16:04
Air Blank		0.000	16:05
Control Test		0.079	16:05
Air Blank		0.000	16:06
Subject Sample	# 1	0.256	16:06
Breath Volume		3.519	16:06
Air Blank		0.000	16:07
Air Blank		0.000	16:09
Subject Sample	# 2	0.247	16:09
Breath Volume		3.851	16:09
Air Blank		0.000	16:10
Control Test		0.078	16:10
Air Blank		0.000	16:11
Diagnostics Check		OK	16:11

Data Download Date/Time: 11/29/2018 10:41

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,CODY,L
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: BROWN,CODY
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: XOCHIQUIQUISQUI

DATE/TIME OF TEST: 11/23/2018 01:42:11

Diagnostics Check		OK	01:35
Air Blank		0.000	01:36
Control Test		0.079	01:36
Air Blank		0.000	01:36
Subject Sample	# 1	0.033	01:37
Breath Volume		1.125	01:37
Air Blank		0.000	01:37
Air Blank		0.000	01:39
Subject Sample	# 2	0.032	01:40
Breath Volume		1.300	01:40
Air Blank		0.000	01:41
Control Test		0.078	01:41
Air Blank		0.000	01:42
Diagnostics Check		OK	01:42

Data Download Date/Time: 11/29/2018 10:41

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,JEREMY,A
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: LEWIS,JEREMY
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: JEAN

DATE/TIME OF TEST: 11/24/2018 04:23:44

Diagnostics Check		OK	04:07	
Air Blank		0.000	04:08	
Control Test		0.079	04:08	
Air Blank		0.000	04:09	
Subject Sample	# 1	0.138	04:12	Volume Not Met
Breath Volume		0.664	04:12	
Air Blank		0.000	04:12	
Air Blank		0.000	04:14	
Subject Sample	# 2	0.144	04:16	
Breath Volume		1.421	04:16	
Air Blank		0.000	04:17	
Air Blank		0.000	04:19	
Subject Sample	# 3	0.138	04:22	Slope Not Level
Breath Volume		1.648	04:22	
Air Blank		0.000	04:22	
Control Test		0.077	04:23	
Air Blank		0.000	04:23	
Diagnostics Check		OK	04:23	

Data Download Date/Time: 11/29/2018 10:41

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,RICHARD,D
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: HALL,RICHARD
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 06:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:15

SUBJECT'S LAST NAME: RODRIGUEZ DE JESUS

DATE/TIME OF TEST: 11/24/2018 07:49:42

Diagnostics Check		OK	07:41
Air Blank		0.000	07:41
Control Test		0.079	07:42
Air Blank		0.000	07:42
Subject Sample	# 1	0.221	07:44
Breath Volume		2.460	07:44
Air Blank		0.000	07:45
Air Blank		0.000	07:47
Subject Sample	# 2	0.223	07:48
Breath Volume		2.500	07:48
Air Blank		0.000	07:48
Control Test		0.077	07:49
Air Blank		0.000	07:49
Diagnostics Check		OK	07:49

Data Download Date/Time: 11/29/2018 10:41

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OJEDA,NELSON,J
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: MCINERNEY,JESSE
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: BENTANCOURT

DATE/TIME OF TEST: 11/25/2018 02:13:12

Diagnostics Check		OK	02:03	
Air Blank		0.000	02:03	
Control Test		0.079	02:04	
Air Blank		0.000	02:04	
Subject Sample	# 1	0.155	02:07	Volume Not Met
Breath Volume		0.542	02:07	
Air Blank		0.000	02:08	
Air Blank		0.000	02:10	
Subject Sample	# 2		02:11	Subject Test Refused
Breath Volume		0.468	02:11	
Air Blank		0.000	02:12	
Control Test		0.078	02:12	
Air Blank		0.000	02:13	
Diagnostics Check		OK	02:13	

Data Download Date/Time: 11/29/2018 10:41

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAX,ERIC,C
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: SERAPHIN,RONALD
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:01

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 11/25/2018 21:23:26

Diagnostics Check		OK	21:12
Air Blank		0.000	21:13
Control Test		0.080	21:13
Air Blank		0.000	21:13
Subject Sample	# 1	0.140	21:16
Breath Volume		2.660	21:16
Air Blank		0.000	21:17
Air Blank		0.000	21:19
Subject Sample	# 2	0.133	21:21
Breath Volume		2.539	21:21
Air Blank		0.000	21:22
Control Test		0.079	21:22
Air Blank		0.000	21:23
Diagnostics Check		OK	21:23

Data Download Date/Time: 11/29/2018 10:41

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: DRAWDY,LAINE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 11/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: BLAZQUEZ

DATE/TIME OF TEST: 11/30/2018 01:10:47

Diagnostics Check		OK	01:04
Air Blank		0.000	01:04
Control Test		0.080	01:04
Air Blank		0.000	01:05
Subject Sample	# 1	0.000	01:06
Breath Volume		2.324	01:06
Air Blank		0.000	01:06
Air Blank		0.000	01:08
Subject Sample	# 2	0.000	01:09
Breath Volume		1.617	01:09
Air Blank		0.000	01:09
Control Test		0.080	01:10
Air Blank		0.000	01:10
Diagnostics Check		OK	01:10

Data Download Date/Time: 12/21/2018 11:31

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,RICHARD,D
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: HALL,RICHARD
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 11/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: BOLDEN

DATE/TIME OF TEST: 11/30/2018 04:09:11

Diagnostics Check		OK	04:03
Air Blank		0.000	04:03
Control Test		0.079	04:03
Air Blank		0.000	04:04
Subject Sample	# 1	0.221	04:04
Breath Volume		3.988	04:04
Air Blank		0.000	04:05
Air Blank		0.000	04:07
Subject Sample	# 2	0.214	04:07
Breath Volume		3.609	04:07
Air Blank		0.000	04:08
Control Test		0.079	04:08
Air Blank		0.000	04:09
Diagnostics Check		OK	04:09

Data Download Date/Time: 12/21/2018 11:31

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000797

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAURES,BRANDYN,K
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BAURES,BRANDYN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 11/27/2018 00:42:32

Diagnostics Check		OK	00:33
Air Blank		0.000	00:34
Control Test		0.082	00:34
Air Blank		0.000	00:35
Subject Sample	# 1	0.117	00:38
Breath Volume		2.738	00:38
Air Blank		0.000	00:38
Air Blank		0.000	00:40
Subject Sample	# 2	0.112	00:41
Breath Volume		2.101	00:41
Air Blank		0.000	00:41
Control Test		0.082	00:41
Air Blank		0.000	00:42
Diagnostics Check		OK	00:42

Data Download Date/Time: 12/03/2018 01:47

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000809

INSTRUMENT REGISTERED TO: DIXIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FULTON,DARIN,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: FULTON,DARIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: GIOIELLI

DATE/TIME OF TEST: 11/06/2018 03:01:09

Diagnostics Check		OK	02:53
Air Blank		0.000	02:54
Control Test		0.080	02:54
Air Blank		0.000	02:55
Subject Sample	# 1	0.213	02:56
Breath Volume		3.144	02:56
Air Blank		0.000	02:57
Air Blank		0.000	02:59
Subject Sample	# 2	0.205	02:59
Breath Volume		2.210	02:59
Air Blank		0.000	03:00
Control Test		0.079	03:00
Air Blank		0.000	03:01
Diagnostics Check		OK	03:01

Data Download Date/Time: 11/13/2018 11:57

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000809

INSTRUMENT REGISTERED TO: DIXIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASH,EARL,C
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MASH,EARL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 11/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1761808A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: DUCOLON

DATE/TIME OF TEST: 11/18/2018 21:58:49

Diagnostics Check		OK	21:51
Air Blank		0.000	21:52
Control Test		0.081	21:52
Air Blank		0.000	21:53
Subject Sample	# 1	0.060	21:54
Breath Volume		2.214	21:54
Air Blank		0.000	21:54
Air Blank		0.000	21:56
Subject Sample	# 2	0.057	21:57
Breath Volume		2.226	21:57
Air Blank		0.000	21:57
Control Test		0.081	21:58
Air Blank		0.000	21:58
Diagnostics Check		OK	21:58

Data Download Date/Time: 12/11/2018 11:43

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000809

INSTRUMENT REGISTERED TO: DIXIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANUS,GLEN,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: GANUS,GLEN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 11/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1761808A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:06

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 11/20/2018 22:32:25

Diagnostics Check		OK	22:26
Air Blank		0.000	22:26
Control Test		0.081	22:27
Air Blank		0.000	22:27
Subject Sample	# 1	0.195	22:28
Breath Volume		2.679	22:28
Air Blank		0.000	22:28
Air Blank		0.000	22:30
Subject Sample	# 2	0.191	22:30
Breath Volume		2.289	22:30
Air Blank		0.000	22:31
Control Test		0.082	22:31
Air Blank		0.000	22:32
Diagnostics Check		OK	22:32

Data Download Date/Time: 12/11/2018 11:43

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000809

INSTRUMENT REGISTERED TO: DIXIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCONNELL,JAIRUS,S
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: FULTON,DARIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 11/13/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1761808A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:23

SUBJECT'S LAST NAME: MCDOWELL

DATE/TIME OF TEST: 11/25/2018 21:52:36

Diagnostics Check		OK	21:44
Air Blank		0.000	21:44
Control Test		0.080	21:44
Air Blank		0.000	21:45
Subject Sample	# 1	0.110	21:48
Breath Volume		2.261	21:48
Air Blank		0.000	21:48
Air Blank		0.000	21:50
Subject Sample	# 2	0.111	21:51
Breath Volume		2.964	21:51
Air Blank		0.000	21:51
Control Test		0.081	21:52
Air Blank		0.000	21:52
Diagnostics Check		OK	21:52

Data Download Date/Time: 12/11/2018 11:43

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000809

INSTRUMENT REGISTERED TO: DIXIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ERNST,WILLIAM,H
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: ANDERSON,SCOTT
ARRESTING OFFICER AGENCY: Dixie County SO
LAST AGENCY INSPECTION: 11/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1761808A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 14:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:25

SUBJECT'S LAST NAME: FUTCH

DATE/TIME OF TEST: 11/28/2018 17:05:50

Diagnostics Check		OK	16:53	
Air Blank		0.000	16:54	
Control Test		0.082	16:54	
Air Blank		0.000	16:54	
Subject Sample	# 1		16:58	No Sample Provided
Breath Volume		0.000	16:58	
Air Blank		0.000	16:58	
Air Blank		0.000	17:00	
Subject Sample	# 2	0.159	17:01	
Breath Volume		2.601	17:01	
Air Blank		0.000	17:01	
Air Blank		0.000	17:03	
Subject Sample	# 3	0.148	17:04	
Breath Volume		2.019	17:04	
Air Blank		0.000	17:04	
Control Test		0.081	17:05	
Air Blank		0.000	17:05	
Diagnostics Check		OK	17:05	

Data Download Date/Time: 12/11/2018 11:43

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLOW,VINESSA,
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: MATA,R
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 11/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: TANNER

DATE/TIME OF TEST: 11/24/2018 22:30:02

Diagnostics Check		OK	22:22
Air Blank		0.000	22:23
Control Test		0.078	22:23
Air Blank		0.000	22:23
Subject Sample	# 1	0.098	22:25
Breath Volume		1.867	22:25
Air Blank		0.000	22:25
Air Blank		0.000	22:27
Subject Sample	# 2	0.086	22:28
Breath Volume		1.375	22:28
Air Blank		0.000	22:29
Control Test		0.075	22:29
Air Blank		0.000	22:29
Diagnostics Check		OK	22:30

Data Download Date/Time: 12/23/2018 09:30

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLOW,VINESSA,
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: ORTIZ,B
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 11/25/2018 04:53:25

Diagnostics Check		OK	04:50	
Air Blank		0.000	04:50	
Control Test		0.079	04:51	
Air Blank		0.000	04:51	
Subject Sample	# 1		04:52	Subject Test Refused
Breath Volume		0.000	04:52	
Air Blank		0.000	04:52	
Control Test		0.079	04:52	
Air Blank		0.000	04:53	
Diagnostics Check		OK	04:53	

Data Download Date/Time: 12/23/2018 09:31

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLOW,VINESSA,
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: BANKASINGH,R
ARRESTING OFFICER AGENCY: Wildwood PD
LAST AGENCY INSPECTION: 11/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: HELTON

DATE/TIME OF TEST: 11/25/2018 23:53:02

Diagnostics Check		OK	23:50	
Air Blank		0.000	23:50	
Control Test		0.080	23:50	
Air Blank		0.000	23:51	
Subject Sample	# 1		23:51	Subject Test Refused
Breath Volume		0.000	23:51	
Air Blank		0.000	23:52	
Control Test		0.079	23:52	
Air Blank		0.000	23:52	
Diagnostics Check		OK	23:53	

Data Download Date/Time: 12/23/2018 09:31

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARSHALL,JARROD,C
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: SCHMIDT,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: RYAN

DATE/TIME OF TEST: 11/27/2018 23:17:38

Diagnostics Check		OK	23:11
Air Blank		0.000	23:11
Control Test		0.080	23:11
Air Blank		0.000	23:12
Subject Sample	# 1	0.108	23:13
Breath Volume		1.945	23:13
Air Blank		0.000	23:13
Air Blank		0.000	23:15
Subject Sample	# 2	0.105	23:16
Breath Volume		1.902	23:16
Air Blank		0.000	23:16
Control Test		0.079	23:17
Air Blank		0.000	23:17
Diagnostics Check		OK	23:17

Data Download Date/Time: 12/23/2018 09:31

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLOW,VINESSA,
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: MATA,R
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 11/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:24

SUBJECT'S LAST NAME: COVIELLO

DATE/TIME OF TEST: 11/29/2018 22:01:28

Diagnostics Check		OK	21:53
Air Blank		0.000	21:53
Control Test		0.079	21:53
Air Blank		0.000	21:54
Subject Sample	# 1	0.254	21:56
Breath Volume		2.183	21:56
Air Blank		0.000	21:57
Air Blank		0.000	21:59
Subject Sample	# 2	0.247	21:59
Breath Volume		1.636	21:59
Air Blank		0.000	22:00
Control Test		0.078	22:00
Air Blank		0.000	22:01
Diagnostics Check		OK	22:01

Data Download Date/Time: 12/23/2018 09:31

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARSHALL,JARROD,C
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: MCCAIG,A
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 10/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: IBARRA

DATE/TIME OF TEST: 11/03/2018 00:11:00

Diagnostics Check		OK	00:04
Air Blank		0.000	00:04
Control Test		0.080	00:04
Air Blank		0.000	00:05
Subject Sample	# 1	0.204	00:05
Breath Volume		1.179	00:05
Air Blank		0.000	00:06
Air Blank		0.000	00:08
Subject Sample	# 2	0.185	00:09
Breath Volume		2.121	00:09
Air Blank		0.000	00:10
Control Test		0.079	00:10
Air Blank		0.000	00:10
Diagnostics Check		OK	00:11

Data Download Date/Time: 11/16/2018 12:15

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEOGHEGAN,ASHLEY,M
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: BERKIHISER,RICHARD
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 10/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: SUDDUTH

DATE/TIME OF TEST: 11/03/2018 17:30:15

Diagnostics Check		OK	17:23	
Air Blank		0.000	17:24	
Control Test		0.080	17:24	
Air Blank		0.000	17:25	
Subject Sample	# 1	0.200	17:25	
Breath Volume		2.785	17:25	
Air Blank		0.000	17:26	
Air Blank		0.000	17:28	
Subject Sample	# 2		17:28	Subject Test Refused
Breath Volume		0.000	17:28	
Air Blank		0.000	17:29	
Control Test		0.080	17:29	
Air Blank		0.000	17:30	
Diagnostics Check		OK	17:30	

Data Download Date/Time: 11/16/2018 12:15

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLOW,VINESSA,
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: WILSON,J
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 10/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: VANDERSNICK

DATE/TIME OF TEST: 11/06/2018 18:34:57

Diagnostics Check		OK	18:26	
Air Blank		0.000	18:26	
Control Test		0.079	18:27	
Air Blank		0.000	18:27	
Subject Sample	# 1	0.217	18:29	
Breath Volume		1.464	18:29	
Air Blank		0.000	18:30	
Air Blank		0.000	18:31	
Subject Sample	# 2	0.210	18:33	
Breath Volume		1.464	18:33	
Air Blank		0.000	18:33	
Control Test		0.074	18:34	Control Outside Tolerance
Air Blank		0.000	18:34	

Data Download Date/Time: 11/16/2018 12:15

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAMBRAZZO,SCOTT,M
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: GUINANNE,J
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 10/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 05:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:45

SUBJECT'S LAST NAME: GAMBRAZZO

DATE/TIME OF TEST: 11/07/2018 07:09:18

Diagnostics Check		OK	07:06	
Air Blank		0.000	07:06	
Control Test		0.079	07:07	
Air Blank		0.000	07:07	
Subject Sample	# 1		07:07	Subject Test Refused
Breath Volume		0.000	07:07	
Air Blank		0.000	07:08	
Control Test		0.079	07:08	
Air Blank		0.000	07:09	
Diagnostics Check		OK	07:09	

Data Download Date/Time: 11/16/2018 12:16

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,JARROD,W
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: COSSEBOOM,MICHAEL
ARRESTING OFFICER AGENCY: Wildwood PD
LAST AGENCY INSPECTION: 10/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: JOUBERT

DATE/TIME OF TEST: 11/07/2018 23:00:46

Diagnostics Check		OK	22:57	
Air Blank		0.000	22:58	
Control Test		0.079	22:58	
Air Blank		0.000	22:59	
Subject Sample	# 1		22:59	Subject Test Refused
Breath Volume		0.000	22:59	
Air Blank		0.000	22:59	
Control Test		0.079	23:00	
Air Blank		0.000	23:00	
Diagnostics Check		OK	23:00	

Data Download Date/Time: 11/16/2018 12:16

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLIS,TERENCE,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: BOUTROS,P
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 10/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 13:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:09

SUBJECT'S LAST NAME: MAGANA BALTAZAR

DATE/TIME OF TEST: 11/08/2018 14:53:59

Diagnostics Check		OK	14:47
Air Blank		0.000	14:48
Control Test		0.079	14:48
Air Blank		0.000	14:49
Subject Sample	# 1	0.000	14:49
Breath Volume		3.152	14:49
Air Blank		0.000	14:50
Air Blank		0.000	14:52
Subject Sample	# 2	0.000	14:52
Breath Volume		3.375	14:52
Air Blank		0.000	14:53
Control Test		0.080	14:53
Air Blank		0.000	14:53
Diagnostics Check		OK	14:53

Data Download Date/Time: 11/16/2018 12:16

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLIS,TERENCE,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: BONURA,M
ARRESTING OFFICER AGENCY: Wildwood PD
LAST AGENCY INSPECTION: 10/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: SIMON JR

DATE/TIME OF TEST: 11/08/2018 18:11:48

Diagnostics Check		OK	18:08	
Air Blank		0.000	18:08	
Control Test		0.079	18:09	
Air Blank		0.000	18:09	
Subject Sample	# 1		18:10	Subject Test Refused
Breath Volume		0.000	18:10	
Air Blank		0.000	18:10	
Control Test		0.079	18:11	
Air Blank		0.000	18:11	
Diagnostics Check		OK	18:11	

Data Download Date/Time: 11/16/2018 12:16

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIVERS,DAWN,T
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: SEARS,TRAVIS
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 10/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 14:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: COTE

DATE/TIME OF TEST: 11/09/2018 15:26:47

Diagnostics Check		OK	15:20
Air Blank		0.000	15:20
Control Test		0.079	15:21
Air Blank		0.000	15:21
Subject Sample	# 1	0.000	15:22
Breath Volume		1.875	15:22
Air Blank		0.000	15:22
Air Blank		0.000	15:24
Subject Sample	# 2	0.000	15:25
Breath Volume		1.742	15:25
Air Blank		0.000	15:25
Control Test		0.080	15:26
Air Blank		0.000	15:26
Diagnostics Check		OK	15:26

Data Download Date/Time: 11/16/2018 12:16

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLOW,VINESSA,
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: SMITH,C
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 10/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: GRUSMARK

DATE/TIME OF TEST: 11/10/2018 01:01:54

Diagnostics Check		OK	00:55
Air Blank		0.000	00:55
Control Test		0.079	00:55
Air Blank		0.000	00:56
Subject Sample	# 1	0.186	00:57
Breath Volume		1.421	00:57
Air Blank		0.000	00:58
Air Blank		0.000	00:59
Subject Sample	# 2	0.182	01:00
Breath Volume		1.636	01:00
Air Blank		0.000	01:00
Control Test		0.078	01:01
Air Blank		0.000	01:01
Diagnostics Check		OK	01:01

Data Download Date/Time: 11/16/2018 12:16

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLOW,VINESSA,
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: BANKASINGH,R
ARRESTING OFFICER AGENCY: Wildwood PD
LAST AGENCY INSPECTION: 10/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: SPURRIER

DATE/TIME OF TEST: 11/10/2018 23:43:02

Diagnostics Check		OK	23:39	
Air Blank		0.000	23:39	
Control Test		0.079	23:39	
Air Blank		0.000	23:40	
Subject Sample	# 1		23:41	Subject Test Refused
Breath Volume		0.000	23:41	
Air Blank		0.000	23:42	
Control Test		0.079	23:42	
Air Blank		0.000	23:42	
Diagnostics Check		OK	23:43	

Data Download Date/Time: 11/16/2018 12:16

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLOW,VINESSA,
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: MATA,R
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 10/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: MCCARTHY

DATE/TIME OF TEST: 11/16/2018 01:15:02

Diagnostics Check		OK	01:06
Air Blank		0.000	01:07
Control Test		0.080	01:07
Air Blank		0.000	01:08
Subject Sample	# 1	0.127	01:09
Breath Volume		1.179	01:09
Air Blank		0.000	01:10
Air Blank		0.000	01:12
Subject Sample	# 2	0.124	01:13
Breath Volume		1.269	01:13
Air Blank		0.000	01:14
Control Test		0.078	01:14
Air Blank		0.000	01:14
Diagnostics Check		OK	01:15

Data Download Date/Time: 11/16/2018 12:16

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ECHEVARRIA,LISA,V
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: TAYLOR,TROY
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 13:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:20

SUBJECT'S LAST NAME: FLYNN

DATE/TIME OF TEST: 11/02/2018 14:47:29

Diagnostics Check		OK	14:44	
Air Blank		0.000	14:45	
Control Test		0.078	14:45	
Air Blank		0.000	14:45	
Subject Sample	# 1		14:46	Subject Test Refused
Breath Volume		0.000	14:46	
Air Blank		0.000	14:46	
Control Test		0.078	14:46	
Air Blank		0.000	14:47	
Diagnostics Check		OK	14:47	

Data Download Date/Time: 11/05/2018 10:08

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLOYD,DANIEL,Z
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: HOWARD,DEP
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:18

SUBJECT'S LAST NAME: BOLES

DATE/TIME OF TEST: 11/03/2018 21:41:15

Diagnostics Check		OK	21:38	
Air Blank		0.000	21:38	
Control Test		0.079	21:39	
Air Blank		0.000	21:39	
Subject Sample	# 1		21:39	Subject Test Refused
Breath Volume		0.000	21:39	
Air Blank		0.000	21:40	
Control Test		0.078	21:40	
Air Blank		0.000	21:41	
Diagnostics Check		OK	21:41	

Data Download Date/Time: 11/05/2018 10:08

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLOYD,DANIEL,Z
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: WELLS,TROOPER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: PORCELLI

DATE/TIME OF TEST: 11/03/2018 23:17:26

Diagnostics Check		OK	23:10
Air Blank		0.000	23:11
Control Test		0.079	23:11
Air Blank		0.000	23:11
Subject Sample	# 1	0.000	23:12
Breath Volume		3.601	23:12
Air Blank		0.000	23:13
Air Blank		0.000	23:15
Subject Sample	# 2	0.000	23:16
Breath Volume		3.925	23:16
Air Blank		0.000	23:16
Control Test		0.079	23:16
Air Blank		0.000	23:17
Diagnostics Check		OK	23:17

Data Download Date/Time: 11/05/2018 10:08

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ECHEVARRIA,LISA,V
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: DOCTOR,JOSH
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:05

SUBJECT'S LAST NAME: MAIER

DATE/TIME OF TEST: 11/04/2018 17:40:44

Diagnostics Check		OK	17:35	
Air Blank		0.000	17:35	
Control Test		0.080	17:36	
Air Blank		0.000	17:36	
Subject Sample	# 1		17:39	Subject Test Refused
Breath Volume		0.000	17:39	
Air Blank		0.000	17:39	
Control Test		0.078	17:40	
Air Blank		0.000	17:40	
Diagnostics Check		OK	17:40	

Data Download Date/Time: 11/05/2018 10:08

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERMASTER,MARY,E
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: FERNANDEZ,C
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: GOMES

DATE/TIME OF TEST: 11/08/2018 03:27:19

Diagnostics Check		OK	03:21
Air Blank		0.000	03:21
Control Test		0.077	03:21
Air Blank		0.000	03:22
Subject Sample	# 1	0.155	03:22
Breath Volume		1.867	03:22
Air Blank		0.000	03:23
Air Blank		0.000	03:25
Subject Sample	# 2	0.156	03:25
Breath Volume		2.191	03:25
Air Blank		0.000	03:26
Control Test		0.078	03:26
Air Blank		0.000	03:27
Diagnostics Check		OK	03:27

Data Download Date/Time: 11/15/2018 17:58

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERMASTER,MARY,E
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: FERNANDEZ,C
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: GOMES

DATE/TIME OF TEST: 11/08/2018 04:11:44

Diagnostics Check		OK	04:05
Air Blank		0.000	04:06
Control Test		0.078	04:06
Air Blank		0.000	04:06
Subject Sample	# 1	0.142	04:07
Breath Volume		1.460	04:07
Air Blank		0.000	04:07
Air Blank		0.000	04:09
Subject Sample	# 2	0.137	04:10
Breath Volume		1.871	04:10
Air Blank		0.000	04:10
Control Test		0.077	04:11
Air Blank		0.000	04:11
Diagnostics Check		OK	04:11

Data Download Date/Time: 11/15/2018 17:58

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHOETTNER,ADRIAN,J
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: HORTON,T
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:56

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 11/08/2018 20:30:25

Diagnostics Check		OK	20:21
Air Blank		0.000	20:22
Control Test		0.078	20:22
Air Blank		0.000	20:23
Subject Sample	# 1	0.146	20:24
Breath Volume		1.769	20:24
Air Blank		0.000	20:24
Air Blank		0.000	20:26
Subject Sample	# 2	0.135	20:28
Breath Volume		1.671	20:28
Air Blank		0.000	20:29
Control Test		0.078	20:29
Air Blank		0.000	20:30
Diagnostics Check		OK	20:30

Data Download Date/Time: 11/15/2018 17:58

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHOETTNER,ADRIAN,J
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: KJEKSTAD,E
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: TRAMMELL

DATE/TIME OF TEST: 11/11/2018 02:09:41

Diagnostics Check		OK	02:06	
Air Blank		0.000	02:07	
Control Test		0.079	02:07	
Air Blank		0.000	02:08	
Subject Sample	# 1		02:08	Subject Test Refused
Breath Volume		0.000	02:08	
Air Blank		0.000	02:08	
Control Test		0.078	02:09	
Air Blank		0.000	02:09	
Diagnostics Check		OK	02:09	

Data Download Date/Time: 11/15/2018 17:59

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHOETTNER,ADRIAN,J
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: ORTIZ,B
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: TROIDE

DATE/TIME OF TEST: 11/11/2018 02:41:37

Diagnostics Check		OK	02:38	
Air Blank		0.000	02:39	
Control Test		0.078	02:39	
Air Blank		0.000	02:40	
Subject Sample	# 1		02:40	Subject Test Refused
Breath Volume		0.000	02:40	
Air Blank		0.000	02:40	
Control Test		0.079	02:41	
Air Blank		0.000	02:41	
Diagnostics Check		OK	02:41	

Data Download Date/Time: 11/15/2018 17:59

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ECHEVARRIA,LISA,V
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: WIGGLESWORTH,CHARLES
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 11/14/2018 03:59:08

Diagnostics Check		OK	03:46	
Air Blank		0.000	03:47	
Control Test		0.079	03:47	
Air Blank		0.000	03:48	
Subject Sample	# 1		03:51	No Sample Provided
Breath Volume		0.000	03:51	
Air Blank		0.000	03:51	
Air Blank		0.000	03:53	
Subject Sample	# 2	0.171	03:54	
Breath Volume		2.621	03:54	
Air Blank		0.000	03:55	
Air Blank		0.000	03:57	
Subject Sample	# 3	0.164	03:57	
Breath Volume		2.332	03:57	
Air Blank		0.000	03:58	
Control Test		0.078	03:58	
Air Blank		0.000	03:59	
Diagnostics Check		OK	03:59	

Data Download Date/Time: 11/15/2018 17:59

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ECHEVARRIA,LISA,V
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: BANG,ROBERT
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 11/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: DORNBLUT

DATE/TIME OF TEST: 11/17/2018 00:22:45

Diagnostics Check		OK	00:14
Air Blank		0.000	00:15
Control Test		0.079	00:15
Air Blank		0.000	00:16
Subject Sample	# 1	0.080	00:17
Breath Volume		1.523	00:17
Air Blank		0.000	00:18
Air Blank		0.000	00:20
Subject Sample	# 2	0.080	00:21
Breath Volume		1.957	00:21
Air Blank		0.000	00:21
Control Test		0.078	00:22
Air Blank		0.000	00:22
Diagnostics Check		OK	00:22

Data Download Date/Time: 12/31/2018 10:10

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ECHEVARRIA,LISA,V
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: WIGGLESWORTH,CHARLES
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 11/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: MCKIGHT

DATE/TIME OF TEST: 11/18/2018 01:46:43

Diagnostics Check		OK	01:43	
Air Blank		0.000	01:44	
Control Test		0.078	01:44	
Air Blank		0.000	01:45	
Subject Sample	# 1		01:45	Subject Test Refused
Breath Volume		0.000	01:45	
Air Blank		0.000	01:45	
Control Test		0.078	01:46	
Air Blank		0.000	01:46	
Diagnostics Check		OK	01:46	

Data Download Date/Time: 12/31/2018 10:10

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHOETTNER,ADRIAN,J
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: BANG,B
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 11/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: CARPENTER

DATE/TIME OF TEST: 11/22/2018 01:44:26

Diagnostics Check		OK	01:37
Air Blank		0.000	01:38
Control Test		0.078	01:38
Air Blank		0.000	01:38
Subject Sample	# 1	0.218	01:39
Breath Volume		2.519	01:39
Air Blank		0.000	01:40
Air Blank		0.000	01:42
Subject Sample	# 2	0.201	01:42
Breath Volume		2.683	01:42
Air Blank		0.000	01:43
Control Test		0.078	01:43
Air Blank		0.000	01:44
Diagnostics Check		OK	01:44

Data Download Date/Time: 12/31/2018 10:10

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHOETTNER,ADRIAN,J
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: WARD,S
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: HARWELL

DATE/TIME OF TEST: 11/22/2018 03:29:25

Diagnostics Check		OK	03:22
Air Blank		0.000	03:23
Control Test		0.079	03:23
Air Blank		0.000	03:23
Subject Sample	# 1	0.186	03:24
Breath Volume		2.960	03:24
Air Blank		0.000	03:25
Air Blank		0.000	03:27
Subject Sample	# 2	0.198	03:27
Breath Volume		1.671	03:27
Air Blank		0.000	03:28
Control Test		0.079	03:28
Air Blank		0.000	03:29
Diagnostics Check		OK	03:29

Data Download Date/Time: 12/31/2018 10:10

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLOYD,DANIEL,Z
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: TAYLOR,DEP
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 11/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 14:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:50

SUBJECT'S LAST NAME: MABREY

DATE/TIME OF TEST: 11/22/2018 15:18:21

Diagnostics Check		OK	15:10
Air Blank		0.000	15:10
Control Test		0.078	15:11
Air Blank		0.000	15:11
Subject Sample	# 1	0.199	15:13
Breath Volume		1.707	15:13
Air Blank		0.000	15:14
Air Blank		0.000	15:16
Subject Sample	# 2	0.193	15:16
Breath Volume		1.718	15:16
Air Blank		0.000	15:17
Control Test		0.077	15:17
Air Blank		0.000	15:18
Diagnostics Check		OK	15:18

Data Download Date/Time: 12/31/2018 10:10

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ECHEVARRIA,LISA,V
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: ALEXANDER,D
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: JAMES

DATE/TIME OF TEST: 11/22/2018 22:15:12

Diagnostics Check		OK	22:09	
Air Blank		0.000	22:09	
Control Test		0.078	22:10	
Air Blank		0.000	22:10	
Subject Sample	# 1		22:13	Subject Test Refused
Breath Volume		0.773	22:13	
Air Blank		0.000	22:14	
Control Test		0.079	22:14	
Air Blank		0.000	22:15	
Diagnostics Check		OK	22:15	

Data Download Date/Time: 12/31/2018 10:10

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLOYD,DANIEL,Z
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: SHEDEKER,DEP
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 11/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:36

SUBJECT'S LAST NAME: TREACY

DATE/TIME OF TEST: 11/23/2018 18:03:24

Diagnostics Check		OK	17:56
Air Blank		0.000	17:57
Control Test		0.078	17:57
Air Blank		0.000	17:58
Subject Sample	# 1	0.000	17:58
Breath Volume		1.894	17:58
Air Blank		0.000	17:59
Air Blank		0.000	18:01
Subject Sample	# 2	0.000	18:02
Breath Volume		1.984	18:02
Air Blank		0.000	18:02
Control Test		0.079	18:02
Air Blank		0.000	18:03
Diagnostics Check		OK	18:03

Data Download Date/Time: 12/31/2018 10:10

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHOETTNER,ADRIAN,J
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: BARR,J
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 11/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: BLANTON

DATE/TIME OF TEST: 11/25/2018 01:39:23

Diagnostics Check		OK	01:36	
Air Blank		0.000	01:37	
Control Test		0.078	01:37	
Air Blank		0.000	01:37	
Subject Sample	# 1		01:38	Subject Test Refused
Breath Volume		0.000	01:38	
Air Blank		0.000	01:38	
Control Test		0.078	01:38	
Air Blank		0.000	01:39	
Diagnostics Check		OK	01:39	

Data Download Date/Time: 12/31/2018 10:10

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHOETTNER,ADRIAN,J
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: HORTON,T
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 11/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: TOOMEY

DATE/TIME OF TEST: 11/26/2018 21:42:23

Diagnostics Check		OK	21:39	
Air Blank		0.000	21:39	
Control Test		0.077	21:40	
Air Blank		0.000	21:40	
Subject Sample	# 1		21:41	Subject Test Refused
Breath Volume		0.000	21:41	
Air Blank		0.000	21:41	
Control Test		0.078	21:41	
Air Blank		0.000	21:42	
Diagnostics Check		OK	21:42	

Data Download Date/Time: 12/31/2018 10:10

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLOYD,DANIEL,Z
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: BARD,TROOPER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 14:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:38

SUBJECT'S LAST NAME: FROST

DATE/TIME OF TEST: 11/27/2018 16:06:27

Diagnostics Check		OK	15:59
Air Blank		0.000	15:59
Control Test		0.078	15:59
Air Blank		0.000	16:00
Subject Sample	# 1	0.279	16:01
Breath Volume		1.230	16:01
Air Blank		0.000	16:02
Air Blank		0.000	16:04
Subject Sample	# 2	0.276	16:04
Breath Volume		1.832	16:04
Air Blank		0.000	16:05
Control Test		0.078	16:05
Air Blank		0.000	16:06
Diagnostics Check		OK	16:06

Data Download Date/Time: 12/31/2018 10:10

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHESSHER,SANDRA,W
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: PITCHER,D
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 10/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 11/02/2018 03:17:17

Diagnostics Check		OK	03:11
Air Blank		0.000	03:11
Control Test		0.080	03:11
Air Blank		0.000	03:12
Subject Sample	# 1	0.123	03:12
Breath Volume		3.562	03:12
Air Blank		0.000	03:13
Air Blank		0.000	03:15
Subject Sample	# 2	0.127	03:15
Breath Volume		3.488	03:15
Air Blank		0.000	03:16
Control Test		0.079	03:16
Air Blank		0.000	03:17
Diagnostics Check		OK	03:17

Data Download Date/Time: 11/15/2018 21:08

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHESSHER,SANDRA,W
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: BINDER,B
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 10/18/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: FLANNERY

DATE/TIME OF TEST: 11/10/2018 01:54:15

Diagnostics Check		OK	01:46
Air Blank		0.000	01:47
Control Test		0.077	01:47
Air Blank		0.000	01:47
Subject Sample	# 1	0.173	01:48
Breath Volume		3.925	01:48
Air Blank		0.000	01:49
Air Blank		0.000	01:50
Subject Sample	# 2	0.165	01:52
Breath Volume		4.089	01:52
Air Blank		0.000	01:53
Control Test		0.076	01:53
Air Blank		0.000	01:54
Diagnostics Check		OK	01:54

Data Download Date/Time: 11/15/2018 21:06

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITCHER,DARRYL,P
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CHESSHER,SANDI
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 10/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: ACKBAR

DATE/TIME OF TEST: 11/11/2018 01:04:42

Diagnostics Check		OK	00:58
Air Blank		0.000	00:58
Control Test		0.076	00:59
Air Blank		0.000	00:59
Subject Sample	# 1	0.116	01:00
Breath Volume		3.398	01:00
Air Blank		0.000	01:00
Air Blank		0.000	01:02
Subject Sample	# 2	0.116	01:03
Breath Volume		3.253	01:03
Air Blank		0.000	01:03
Control Test		0.077	01:04
Air Blank		0.000	01:04
Diagnostics Check		OK	01:04

Data Download Date/Time: 11/15/2018 21:06

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITCHER,DARRYL,P
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: PITCHER,DARRYL
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 11/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: DAUSA

DATE/TIME OF TEST: 11/17/2018 01:21:31

Diagnostics Check		OK	01:18	
Air Blank		0.000	01:18	
Control Test		0.078	01:19	
Air Blank		0.000	01:19	
Subject Sample	# 1		01:20	Subject Test Refused
Breath Volume		0.000	01:20	
Air Blank		0.000	01:20	
Control Test		0.078	01:20	
Air Blank		0.000	01:21	
Diagnostics Check		OK	01:21	

Data Download Date/Time: 12/06/2018 01:01

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINDER,BRONSON,M
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: BINDER,BRONSON
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 11/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: EARL

DATE/TIME OF TEST: 11/24/2018 00:56:42

Diagnostics Check		OK	00:50
Air Blank		0.000	00:50
Control Test		0.076	00:51
Air Blank		0.000	00:51
Subject Sample	# 1	0.166	00:52
Breath Volume		3.566	00:52
Air Blank		0.000	00:52
Air Blank		0.000	00:54
Subject Sample	# 2	0.160	00:55
Breath Volume		2.042	00:55
Air Blank		0.000	00:55
Control Test		0.078	00:56
Air Blank		0.000	00:56
Diagnostics Check		OK	00:56

Data Download Date/Time: 12/06/2018 01:01

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITCHER,DARRYL,P
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: BINDER,BRONSON
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 11/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: FOX-SCHMIDT

DATE/TIME OF TEST: 11/24/2018 22:56:05

Diagnostics Check		OK	22:49
Air Blank		0.000	22:50
Control Test		0.077	22:50
Air Blank		0.000	22:50
Subject Sample	# 1	0.272	22:51
Breath Volume		1.992	22:51
Air Blank		0.000	22:52
Air Blank		0.000	22:53
Subject Sample	# 2	0.262	22:54
Breath Volume		1.968	22:54
Air Blank		0.000	22:55
Control Test		0.077	22:55
Air Blank		0.000	22:56
Diagnostics Check		OK	22:56

Data Download Date/Time: 12/06/2018 01:01

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEVENS,JEREMY,L
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: STEVENS,JEREMY
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 10/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A2
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: STEPHAN

DATE/TIME OF TEST: 11/03/2018 03:52:52

Diagnostics Check		OK	03:50	
Air Blank		0.000	03:50	
Control Test		0.080	03:50	
Air Blank		0.000	03:51	
Subject Sample	# 1		03:51	Subject Test Refused
Breath Volume		0.000	03:51	
Air Blank		0.000	03:51	
Control Test		0.079	03:52	
Air Blank		0.000	03:52	
Diagnostics Check		OK	03:52	

Data Download Date/Time: 11/25/2018 21:42

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: BRUNO

DATE/TIME OF TEST: 11/03/2018 04:24:47

Diagnostics Check		OK	04:18
Air Blank		0.000	04:18
Control Test		0.079	04:18
Air Blank		0.000	04:19
Subject Sample	# 1	0.072	04:20
Breath Volume		2.335	04:20
Air Blank		0.000	04:20
Air Blank		0.000	04:22
Subject Sample	# 2	0.073	04:23
Breath Volume		2.902	04:23
Air Blank		0.000	04:23
Control Test		0.079	04:24
Air Blank		0.000	04:24
Diagnostics Check		OK	04:24

Data Download Date/Time: 12/18/2018 19:30

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HARVESTER,W
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: RODRIGUEZ MENDOZA

DATE/TIME OF TEST: 11/04/2018 02:13:21

Diagnostics Check		OK	02:06
Air Blank		0.000	02:06
Control Test		0.080	02:06
Air Blank		0.000	02:07
Subject Sample	# 1	0.136	02:08
Breath Volume		1.328	02:08
Air Blank		0.000	02:09
Air Blank		0.000	02:10
Subject Sample	# 2	0.141	02:11
Breath Volume		1.925	02:11
Air Blank		0.000	02:12
Control Test		0.079	02:12
Air Blank		0.000	02:13
Diagnostics Check		OK	02:13

Data Download Date/Time: 12/18/2018 19:30

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: HARRISON

DATE/TIME OF TEST: 11/05/2018 01:56:16

Diagnostics Check	OK	01:54	
Air Blank	0.000	01:54	
Control Test	0.080	01:55	
Air Blank	0.000	01:55	
		01:55	Improper Sample
Air Blank	0.000	01:56	

Data Download Date/Time: 12/18/2018 19:30

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: HARRISON

DATE/TIME OF TEST: 11/05/2018 02:04:19

Diagnostics Check		OK	01:57
Air Blank		0.000	01:58
Control Test		0.079	01:58
Air Blank		0.000	01:59
Subject Sample	# 1	0.088	01:59
Breath Volume		1.312	01:59
Air Blank		0.000	02:00
Air Blank		0.000	02:02
Subject Sample	# 2	0.105	02:02
Breath Volume		2.390	02:02
Air Blank		0.000	02:03
Control Test		0.079	02:03
Air Blank		0.000	02:04
Diagnostics Check		OK	02:04

Data Download Date/Time: 12/18/2018 19:30

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: PUEBLA ORTEGA

DATE/TIME OF TEST: 11/05/2018 05:21:41

Diagnostics Check		OK	05:10
Air Blank		0.000	05:11
Control Test		0.079	05:11
Air Blank		0.000	05:12
Subject Sample	# 1	0.201	05:13
Breath Volume		1.968	05:13
Air Blank		0.000	05:14
Air Blank		0.000	05:16
Subject Sample	# 2	0.179	05:16
Breath Volume		1.160	05:16
Air Blank		0.000	05:17
Air Blank		0.000	05:19
Subject Sample	# 3	0.198	05:20
Breath Volume		2.414	05:20
Air Blank		0.000	05:20
Control Test		0.079	05:21
Air Blank		0.000	05:21
Diagnostics Check		OK	05:21

Data Download Date/Time: 12/18/2018 19:30

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: THORNE,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: ANTELL

DATE/TIME OF TEST: 11/06/2018 01:39:38

Diagnostics Check		OK	01:33
Air Blank		0.000	01:33
Control Test		0.079	01:34
Air Blank		0.000	01:34
Subject Sample	# 1	0.132	01:35
Breath Volume		2.957	01:35
Air Blank		0.000	01:35
Air Blank		0.000	01:37
Subject Sample	# 2	0.121	01:38
Breath Volume		1.691	01:38
Air Blank		0.000	01:38
Control Test		0.079	01:39
Air Blank		0.000	01:39
Diagnostics Check		OK	01:39

Data Download Date/Time: 12/18/2018 19:30

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LYON,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: MCCLURKIN

DATE/TIME OF TEST: 11/09/2018 03:39:47

Diagnostics Check		OK	03:31
Air Blank		0.000	03:32
Control Test		0.080	03:32
Air Blank		0.000	03:32
Subject Sample	# 1	0.113	03:35
Breath Volume		1.558	03:35
Air Blank		0.000	03:35
Air Blank		0.000	03:37
Subject Sample	# 2	0.112	03:38
Breath Volume		1.996	03:38
Air Blank		0.000	03:38
Control Test		0.079	03:39
Air Blank		0.000	03:39
Diagnostics Check		OK	03:39

Data Download Date/Time: 12/18/2018 19:30

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LYON,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: RICHEY

DATE/TIME OF TEST: 11/10/2018 00:53:17

Diagnostics Check		OK	00:46
Air Blank		0.000	00:47
Control Test		0.080	00:47
Air Blank		0.000	00:48
Subject Sample	# 1	0.163	00:48
Breath Volume		1.468	00:48
Air Blank		0.000	00:49
Air Blank		0.000	00:51
Subject Sample	# 2	0.175	00:51
Breath Volume		1.652	00:51
Air Blank		0.000	00:52
Control Test		0.079	00:52
Air Blank		0.000	00:53
Diagnostics Check		OK	00:53

Data Download Date/Time: 12/18/2018 19:30

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MOREJON,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: TANNER

DATE/TIME OF TEST: 11/11/2018 00:52:17

Diagnostics Check		OK	00:44
Air Blank		0.000	00:45
Control Test		0.080	00:45
Air Blank		0.000	00:46
Subject Sample	# 1	0.000	00:47
Breath Volume		2.667	00:47
Air Blank		0.000	00:48
Air Blank		0.000	00:50
Subject Sample	# 2	0.000	00:50
Breath Volume		1.292	00:50
Air Blank		0.000	00:51
Control Test		0.079	00:51
Air Blank		0.000	00:52
Diagnostics Check		OK	00:52

Data Download Date/Time: 12/18/2018 19:30

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MACIAS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:41

SUBJECT'S LAST NAME: VELASQUEZ VELASQUEZ

DATE/TIME OF TEST: 11/11/2018 21:07:23

Diagnostics Check		OK	21:01
Air Blank		0.000	21:01
Control Test		0.080	21:01
Air Blank		0.000	21:02
Subject Sample	# 1	0.166	21:02
Breath Volume		1.777	21:02
Air Blank		0.000	21:03
Air Blank		0.000	21:05
Subject Sample	# 2	0.174	21:05
Breath Volume		2.445	21:05
Air Blank		0.000	21:06
Control Test		0.079	21:06
Air Blank		0.000	21:07
Diagnostics Check		OK	21:07

Data Download Date/Time: 12/18/2018 19:30

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: SHAW JR

DATE/TIME OF TEST: 11/12/2018 21:29:59

Diagnostics Check		OK	21:21
Air Blank		0.000	21:22
Control Test		0.080	21:22
Air Blank		0.000	21:22
Subject Sample	# 1	0.226	21:25
Breath Volume		1.343	21:25
Air Blank		0.000	21:25
Air Blank		0.000	21:27
Subject Sample	# 2	0.231	21:28
Breath Volume		1.460	21:28
Air Blank		0.000	21:29
Control Test		0.080	21:29
Air Blank		0.000	21:29
Diagnostics Check		OK	21:29

Data Download Date/Time: 12/18/2018 19:30

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: NIEBLAS

DATE/TIME OF TEST: 11/14/2018 04:39:15

Diagnostics Check		OK	04:32
Air Blank		0.000	04:33
Control Test		0.080	04:33
Air Blank		0.000	04:34
Subject Sample	# 1	0.145	04:34
Breath Volume		2.171	04:34
Air Blank		0.000	04:35
Air Blank		0.000	04:37
Subject Sample	# 2	0.139	04:37
Breath Volume		2.597	04:37
Air Blank		0.000	04:38
Control Test		0.079	04:38
Air Blank		0.000	04:39
Diagnostics Check		OK	04:39

Data Download Date/Time: 12/18/2018 19:31

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:54

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 11/16/2018 05:21:45

Diagnostics Check		OK	05:14
Air Blank		0.000	05:15
Control Test		0.080	05:15
Air Blank		0.000	05:16
Subject Sample	# 1	0.098	05:16
Breath Volume		2.972	05:16
Air Blank		0.000	05:17
Air Blank		0.000	05:19
Subject Sample	# 2	0.102	05:20
Breath Volume		3.113	05:20
Air Blank		0.000	05:20
Control Test		0.080	05:21
Air Blank		0.000	05:21
Diagnostics Check		OK	05:21

Data Download Date/Time: 12/18/2018 19:31

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TIBURCIO,I
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: WHITNEY

DATE/TIME OF TEST: 11/17/2018 02:49:28

Diagnostics Check		OK	02:43
Air Blank		0.000	02:43
Control Test		0.081	02:44
Air Blank		0.000	02:44
Subject Sample	# 1	0.070	02:45
Breath Volume		2.074	02:45
Air Blank		0.000	02:45
Air Blank		0.000	02:47
Subject Sample	# 2	0.070	02:48
Breath Volume		1.667	02:48
Air Blank		0.000	02:48
Control Test		0.080	02:48
Air Blank		0.000	02:49
Diagnostics Check		OK	02:49

Data Download Date/Time: 12/18/2018 19:31

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LYON,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: TYNER

DATE/TIME OF TEST: 11/17/2018 03:51:21

Diagnostics Check		OK	03:44
Air Blank		0.000	03:44
Control Test		0.080	03:44
Air Blank		0.000	03:45
Subject Sample	# 1	0.209	03:46
Breath Volume		4.046	03:46
Air Blank		0.000	03:47
Air Blank		0.000	03:49
Subject Sample	# 2	0.209	03:49
Breath Volume		3.796	03:49
Air Blank		0.000	03:50
Control Test		0.080	03:50
Air Blank		0.000	03:51
Diagnostics Check		OK	03:51

Data Download Date/Time: 12/18/2018 19:31

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 11/18/2018 03:15:59

Diagnostics Check		OK	03:09
Air Blank		0.000	03:10
Control Test		0.080	03:10
Air Blank		0.000	03:10
Subject Sample	# 1	0.147	03:11
Breath Volume		3.042	03:11
Air Blank		0.000	03:12
Air Blank		0.000	03:14
Subject Sample	# 2	0.149	03:14
Breath Volume		2.992	03:14
Air Blank		0.000	03:15
Control Test		0.080	03:15
Air Blank		0.000	03:15
Diagnostics Check		OK	03:15

Data Download Date/Time: 12/18/2018 19:31

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DELORENZO,D
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: CASAS CABRERA

DATE/TIME OF TEST: 11/18/2018 04:31:45

Diagnostics Check		OK	04:25
Air Blank		0.000	04:25
Control Test		0.080	04:26
Air Blank		0.000	04:26
Subject Sample	# 1	0.190	04:27
Breath Volume		1.894	04:27
Air Blank		0.000	04:27
Air Blank		0.000	04:29
Subject Sample	# 2	0.197	04:30
Breath Volume		2.527	04:30
Air Blank		0.000	04:30
Control Test		0.079	04:31
Air Blank		0.000	04:31
Diagnostics Check		OK	04:31

Data Download Date/Time: 12/18/2018 19:31

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DE JESUS,V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:51

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 11/20/2018 00:22:16

Diagnostics Check		OK	00:11
Air Blank		0.000	00:11
Control Test		0.080	00:12
Air Blank		0.000	00:12
Subject Sample	# 1	0.119	00:14
Breath Volume		1.113	00:14
Air Blank		0.000	00:14
Air Blank		0.000	00:16
Subject Sample	# 2	0.143	00:17
Breath Volume		2.195	00:17
Air Blank		0.000	00:17
Air Blank		0.000	00:19
Subject Sample	# 3	0.149	00:20
Breath Volume		2.113	00:20
Air Blank		0.000	00:21
Control Test		0.080	00:21
Air Blank		0.000	00:22
Diagnostics Check		OK	00:22

Data Download Date/Time: 12/18/2018 19:31

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIMPKINS,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: MOYA ACUNA

DATE/TIME OF TEST: 11/21/2018 22:35:47

Diagnostics Check		OK	22:28
Air Blank		0.000	22:29
Control Test		0.081	22:29
Air Blank		0.000	22:30
Subject Sample	# 1	0.238	22:31
Breath Volume		2.003	22:31
Air Blank		0.000	22:31
Air Blank		0.000	22:33
Subject Sample	# 2	0.244	22:34
Breath Volume		2.531	22:34
Air Blank		0.000	22:34
Control Test		0.079	22:35
Air Blank		0.000	22:35
Diagnostics Check		OK	22:35

Data Download Date/Time: 12/18/2018 19:31

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: BRANTLEY

DATE/TIME OF TEST: 11/21/2018 23:02:32

Diagnostics Check		OK	22:55
Air Blank		0.000	22:56
Control Test		0.080	22:56
Air Blank		0.000	22:57
Subject Sample	# 1	0.115	22:57
Breath Volume		1.601	22:57
Air Blank		0.000	22:58
Air Blank		0.000	23:00
Subject Sample	# 2	0.115	23:01
Breath Volume		1.472	23:01
Air Blank		0.000	23:01
Control Test		0.080	23:01
Air Blank		0.000	23:02
Diagnostics Check		OK	23:02

Data Download Date/Time: 12/18/2018 19:31

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SCHAEFER,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: BARTOLON PEREZ

DATE/TIME OF TEST: 11/22/2018 01:26:23

Diagnostics Check		OK	01:18
Air Blank		0.000	01:18
Control Test		0.080	01:18
Air Blank		0.000	01:19
Subject Sample	# 1	0.195	01:21
Breath Volume		1.570	01:21
Air Blank		0.000	01:22
Air Blank		0.000	01:24
Subject Sample	# 2	0.198	01:24
Breath Volume		1.527	01:24
Air Blank		0.000	01:25
Control Test		0.080	01:25
Air Blank		0.000	01:26
Diagnostics Check		OK	01:26

Data Download Date/Time: 12/18/2018 19:31

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,REBECCA,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DRISCOLL,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: WEBB

DATE/TIME OF TEST: 11/23/2018 00:43:24

Diagnostics Check		OK	00:36
Air Blank		0.000	00:36
Control Test		0.080	00:37
Air Blank		0.000	00:37
Subject Sample	# 1	0.116	00:38
Breath Volume		2.855	00:38
Air Blank		0.000	00:39
Air Blank		0.000	00:41
Subject Sample	# 2	0.119	00:41
Breath Volume		2.875	00:41
Air Blank		0.000	00:42
Control Test		0.080	00:42
Air Blank		0.000	00:43
Diagnostics Check		OK	00:43

Data Download Date/Time: 12/18/2018 19:31

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DE JESUS,V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: CRUZ RODRIGUEZ

DATE/TIME OF TEST: 11/25/2018 23:32:44

Diagnostics Check		OK	23:26
Air Blank		0.000	23:26
Control Test		0.079	23:27
Air Blank		0.000	23:27
Subject Sample	# 1	0.121	23:28
Breath Volume		2.347	23:28
Air Blank		0.000	23:29
Air Blank		0.000	23:30
Subject Sample	# 2	0.115	23:31
Breath Volume		1.128	23:31
Air Blank		0.000	23:31
Control Test		0.079	23:32
Air Blank		0.000	23:32
Diagnostics Check		OK	23:32

Data Download Date/Time: 12/18/2018 19:31

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: WEBER,JEFFREY
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 10:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:54

SUBJECT'S LAST NAME: CHONG

DATE/TIME OF TEST: 11/07/2018 12:22:32

Diagnostics Check		OK	12:15
Air Blank		0.000	12:16
Control Test		0.080	12:16
Air Blank		0.000	12:16
Subject Sample	# 1	0.000	12:17
Breath Volume		2.097	12:17
Air Blank		0.000	12:18
Air Blank		0.000	12:20
Subject Sample	# 2	0.000	12:21
Breath Volume		1.894	12:21
Air Blank		0.000	12:21
Control Test		0.080	12:21
Air Blank		0.000	12:22
Diagnostics Check		OK	12:22

Data Download Date/Time: 12/02/2018 18:57

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,JAMES,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: JONES,JAMES
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 05:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:02

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 11/09/2018 06:13:24

Diagnostics Check		OK	06:06
Air Blank		0.000	06:07
Control Test		0.079	06:07
Air Blank		0.000	06:08
Subject Sample	# 1	0.142	06:08
Breath Volume		1.117	06:08
Air Blank		0.000	06:09
Air Blank		0.000	06:11
Subject Sample	# 2	0.155	06:11
Breath Volume		2.152	06:11
Air Blank		0.000	06:12
Control Test		0.079	06:12
Air Blank		0.000	06:13
Diagnostics Check		OK	06:13

Data Download Date/Time: 12/02/2018 18:57

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARAFA,YUSUF,A
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: ARAFA,YUSUF
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: LUNA

DATE/TIME OF TEST: 11/18/2018 05:26:29

Diagnostics Check		OK	05:21	
Air Blank		0.000	05:21	
Control Test		0.081	05:21	
Air Blank		0.000	05:22	
Subject Sample	# 1		05:24	Subject Test Refused
Breath Volume		1.015	05:24	
Air Blank		0.000	05:25	
Control Test		0.081	05:25	
Air Blank		0.000	05:26	
Diagnostics Check		OK	05:26	

Data Download Date/Time: 12/02/2018 18:57

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WURST,JOSHUA,D
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: WURST,JOSHUA
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 09:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:45

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 11/18/2018 10:16:37

Diagnostics Check		OK	10:13	
Air Blank		0.000	10:14	
Control Test		0.081	10:14	
Air Blank		0.000	10:14	
Subject Sample	# 1		10:15	Subject Test Refused
Breath Volume		0.000	10:15	
Air Blank		0.000	10:15	
Control Test		0.081	10:16	
Air Blank		0.000	10:16	
Diagnostics Check		OK	10:16	

Data Download Date/Time: 12/02/2018 18:57

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILLMAN,JAMES,L
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: COLLINS,JAMES
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 14:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:20

SUBJECT'S LAST NAME: DAVISON

DATE/TIME OF TEST: 11/23/2018 15:22:23

Diagnostics Check		OK	15:13
Air Blank		0.000	15:13
Control Test		0.079	15:14
Air Blank		0.000	15:14
Subject Sample	# 1	0.201	15:14
Breath Volume		1.660	15:14
Air Blank		0.000	15:15
Air Blank		0.000	15:17
Subject Sample	# 2	0.229	15:17
Breath Volume		1.316	15:17
Air Blank		0.000	15:18
Air Blank		0.000	15:20
Subject Sample	# 3	0.205	15:20
Breath Volume		2.265	15:20
Air Blank		0.000	15:21
Control Test		0.077	15:21
Air Blank		0.000	15:22
Diagnostics Check		OK	15:22

Data Download Date/Time: 12/02/2018 18:57

Printed: 16 January 2019

COBRA ©1997-2014 CMI, Inc.