

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HAMILTON,J,W

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: MCDONALD,C

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 11/26/2018

Date or Time Changed: Yes

DRY GAS STANDARD LOT #: 22817080A4

DRY GAS EXPIRY DATE: 10/05/2019

VIOLATION CODE: DUI

ARREST TIME: 01:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: KVEREK

DATE/TIME OF TEST: 12/26/2018 01:28:52

Diagnostics Check	OK	01:22
Air Blank	0.000	01:23
Control Test	0.079	01:23
Air Blank	0.000	01:23
Subject Sample # 1	0.109	01:24
Breath Volume	2.527	01:24
Air Blank	0.000	01:25
Air Blank	0.000	01:26
Subject Sample # 2	0.105	01:27
Breath Volume	1.761	01:27
Air Blank	0.000	01:27
Control Test	0.081	01:28
Air Blank	0.000	01:28
Diagnostics Check	OK	01:28

Data Download Date/Time: 12/29/2018 22:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001737

INSTRUMENT REGISTERED TO: SOUTH MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANDA,CHELSY,
BREATH TEST OPERATOR AGENCY: South Miami PD
ARRESTING OFFICER: LANDA,CHELSY
ARRESTING OFFICER AGENCY: South Miami PD
LAST AGENCY INSPECTION: 11/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: COLL

DATE/TIME OF TEST: 12/08/2018 04:06:44

Diagnostics Check	OK	04:03	
Air Blank	0.000	04:04	
Control Test	0.079	04:04	
Air Blank	0.000	04:04	
Subject Sample	# 1	04:05	Subject Test Refused
Breath Volume	0.000	04:05	
Air Blank	0.000	04:05	
Control Test	0.079	04:06	
Air Blank	0.000	04:06	
Diagnostics Check	OK	04:06	

Data Download Date/Time: 12/13/2018 10:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001737

INSTRUMENT REGISTERED TO: SOUTH MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,CHRISTOPHER,E
BREATH TEST OPERATOR AGENCY: South Miami PD
ARRESTING OFFICER: BOBADILLA,JUAN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 12/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: MORENO

DATE/TIME OF TEST: 12/21/2018 23:52:55

Diagnostics Check	OK	23:45
Air Blank	0.000	23:46
Control Test	0.079	23:46
Air Blank	0.000	23:47
Subject Sample # 1	0.267	23:48
Breath Volume	3.078	23:48
Air Blank	0.000	23:49
Air Blank	0.000	23:50
Subject Sample # 2	0.271	23:51
Breath Volume	2.480	23:51
Air Blank	0.000	23:51
Control Test	0.077	23:52
Air Blank	0.000	23:52
Diagnostics Check	OK	23:52

Data Download Date/Time: 01/11/2019 11:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULDERRIG,SEAN,M
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: MULDERRIG,SEAN
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 12/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 14:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:11

SUBJECT'S LAST NAME: WALTENS

DATE/TIME OF TEST: 12/19/2018 15:38:35

Diagnostics Check	OK	15:32
Air Blank	0.000	15:32
Control Test	0.076	15:32
Air Blank	0.000	15:33
Subject Sample # 1	0.000	15:33
Breath Volume	2.859	15:33
Air Blank	0.000	15:34
Air Blank	0.000	15:36
Subject Sample # 2	0.000	15:37
Breath Volume	2.761	15:37
Air Blank	0.000	15:37
Control Test	0.077	15:38
Air Blank	0.000	15:38
Diagnostics Check	OK	15:38

Data Download Date/Time: 01/14/2019 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTLE,CORY,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:10

SUBJECT'S LAST NAME: GILLET

DATE/TIME OF TEST: 12/21/2018 17:36:23

Diagnostics Check	OK	17:30
Air Blank	0.000	17:30
Control Test	0.076	17:31
Air Blank	0.000	17:31
Subject Sample # 1	0.000	17:32
Breath Volume	2.636	17:32
Air Blank	0.000	17:32
Air Blank	0.000	17:34
Subject Sample # 2	0.000	17:35
Breath Volume	1.925	17:35
Air Blank	0.000	17:35
Control Test	0.077	17:35
Air Blank	0.000	17:36
Diagnostics Check	OK	17:36

Data Download Date/Time: 01/14/2019 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTLE,CORY,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: ROBERTS,REBECCA
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 12/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 12/21/2018 22:56:58

Diagnostics Check	OK	22:51
Air Blank	0.000	22:51
Control Test	0.078	22:51
Air Blank	0.000	22:52
Subject Sample # 1	0.000	22:52
Breath Volume	2.492	22:52
Air Blank	0.000	22:53
Air Blank	0.000	22:55
Subject Sample # 2	0.000	22:55
Breath Volume	1.753	22:55
Air Blank	0.000	22:56
Control Test	0.077	22:56
Air Blank	0.000	22:56
Diagnostics Check	OK	22:56

Data Download Date/Time: 01/14/2019 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALMEIDA,CAMILO,J
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: ALMEIDA,CAMILO
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 12/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: TOWNSEND

DATE/TIME OF TEST: 12/22/2018 21:05:05

Diagnostics Check	OK	20:58
Air Blank	0.000	20:59
Control Test	0.078	20:59
Air Blank	0.000	20:59
Subject Sample # 1	0.282	21:00
Breath Volume	1.718	21:00
Air Blank	0.000	21:01
Air Blank	0.000	21:03
Subject Sample # 2	0.285	21:03
Breath Volume	1.394	21:03
Air Blank	0.000	21:04
Control Test	0.078	21:04
Air Blank	0.000	21:05
Diagnostics Check	OK	21:05

Data Download Date/Time: 01/14/2019 10:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ALMEIDA,CAMILO,J

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: ALMEIDA,CAMILO

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 12/19/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 22817080A5

DRY GAS EXPIRY DATE: 10/05/2019

VIOLATION CODE: DUI

ARREST TIME: 03:21

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: SWECKER

DATE/TIME OF TEST: 12/23/2018 04:07:00

Diagnostics Check	OK	04:04	
Air Blank	0.000	04:04	
Control Test	0.078	04:04	
Air Blank	0.000	04:05	
Subject Sample	# 1	04:05	Subject Test Refused
Breath Volume	0.000	04:05	
Air Blank	0.000	04:06	
Control Test	0.078	04:06	
Air Blank	0.000	04:06	
Diagnostics Check	OK	04:07	

Data Download Date/Time: 01/14/2019 10:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002464

INSTRUMENT REGISTERED TO: HIALEAH PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ARROJAS,GABRIEL,L

BREATH TEST OPERATOR AGENCY: Hialeah PD

ARRESTING OFFICER: ELIAS,FELIX

ARRESTING OFFICER AGENCY: Hialeah PD

LAST AGENCY INSPECTION: 11/14/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 712031

DRY GAS EXPIRY DATE: 02/21/2020

VIOLATION CODE: DUI

ARREST TIME: 01:15

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 12/14/2018 03:16:02

Diagnostics Check	OK	03:13	
Air Blank	0.000	03:13	
Control Test	0.079	03:13	
Air Blank	0.000	03:14	
Subject Sample	# 1	03:14	Subject Test Refused
Breath Volume	0.000	03:14	
Air Blank	0.000	03:15	
Control Test	0.078	03:15	
Air Blank	0.000	03:15	
Diagnostics Check	OK	03:16	

Data Download Date/Time: 12/19/2018 10:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002464

INSTRUMENT REGISTERED TO: HIALEAH PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ARROJAS,GABRIEL,L

BREATH TEST OPERATOR AGENCY: Hialeah PD

ARRESTING OFFICER: GARCIA,DANIEL

ARRESTING OFFICER AGENCY: Hialeah PD

LAST AGENCY INSPECTION: 11/14/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 712031

DRY GAS EXPIRY DATE: 02/21/2020

VIOLATION CODE: DUI

ARREST TIME: 03:45

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 12/14/2018 04:38:20

Diagnostics Check	OK	04:35	
Air Blank	0.000	04:35	
Control Test	0.079	04:36	
Air Blank	0.000	04:36	
Subject Sample	# 1	04:36	Subject Test Refused
Breath Volume	0.000	04:36	
Air Blank	0.000	04:37	
Control Test	0.078	04:37	
Air Blank	0.000	04:38	
Diagnostics Check	OK	04:38	

Data Download Date/Time: 12/19/2018 10:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002464

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUNOZ,A,
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: HERNANDEZ,S
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 11/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 14:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:56

SUBJECT'S LAST NAME: ALMARALES

DATE/TIME OF TEST: 12/16/2018 16:31:28

Diagnostics Check	OK	16:25
Air Blank	0.000	16:25
Control Test	0.078	16:25
Air Blank	0.000	16:26
Subject Sample # 1	0.218	16:26
Breath Volume	1.652	16:26
Air Blank	0.000	16:27
Air Blank	0.000	16:29
Subject Sample # 2	0.222	16:29
Breath Volume	1.343	16:29
Air Blank	0.000	16:30
Control Test	0.076	16:30
Air Blank	0.000	16:31
Diagnostics Check	OK	16:31

Data Download Date/Time: 12/19/2018 10:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002987

INSTRUMENT REGISTERED TO: MIAMI GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROMERO,CARLOS,A
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: ROMERO,CARLOS
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 11/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: REYNOLDS

DATE/TIME OF TEST: 12/23/2018 02:58:36

Diagnostics Check	OK	02:55	
Air Blank	0.000	02:56	
Control Test	0.079	02:56	
Air Blank	0.000	02:56	
Subject Sample	# 1	02:57	Subject Test Refused
Breath Volume	0.000	02:57	
Air Blank	0.000	02:57	
Control Test	0.080	02:58	
Air Blank	0.000	02:58	
Diagnostics Check	OK	02:58	

Data Download Date/Time: 12/27/2018 09:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002988

INSTRUMENT REGISTERED TO: MIAMI GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HODGES,TITUS,R
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: HODGES,TITUS
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 11/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: PERSELL

DATE/TIME OF TEST: 12/16/2018 04:58:29

Diagnostics Check	OK	04:49	
Air Blank	0.000	04:50	
Control Test	0.079	04:50	
Air Blank	0.000	04:51	
Subject Sample	# 1	04:54	No Sample Provided
Breath Volume	0.000	04:54	
Air Blank	0.000	04:54	
Air Blank	0.000	04:56	
Subject Sample	# 2	04:57	Subject Test Refused
Breath Volume	0.000	04:57	
Air Blank	0.000	04:57	
Control Test	0.079	04:57	
Air Blank	0.000	04:58	
Diagnostics Check	OK	04:58	

Data Download Date/Time: 12/27/2018 09:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,G,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,G
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 11/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: OLESZEK

DATE/TIME OF TEST: 12/06/2018 02:17:08

Diagnostics Check	OK	02:10
Air Blank	0.000	02:11
Control Test	0.080	02:11
Air Blank	0.000	02:12
Subject Sample # 1	0.131	02:12
Breath Volume	3.039	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample # 2	0.135	02:15
Breath Volume	2.765	02:15
Air Blank	0.000	02:16
Control Test	0.079	02:16
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

Data Download Date/Time: 12/13/2018 09:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAIGE,LEON,I
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: PAIGE,LEON
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: CHARUL SIGURANI

DATE/TIME OF TEST: 12/01/2018 02:50:57

Diagnostics Check	OK	02:42
Air Blank	0.000	02:42
Control Test	0.080	02:43
Air Blank	0.000	02:43
Subject Sample # 1	0.325	02:45
Breath Volume	1.207	02:45
Air Blank	0.000	02:46
Air Blank	0.000	02:47
Subject Sample # 2	0.322	02:49
Breath Volume	2.078	02:49
Air Blank	0.000	02:49
Control Test	0.076	02:50
Air Blank	0.000	02:50
Diagnostics Check	OK	02:50

Data Download Date/Time: 12/27/2018 13:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COSCULLUELA,BRYAN,M
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: STANTON,TARYN
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: BUI
ARREST TIME: 10:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:45

SUBJECT'S LAST NAME: ARGUELLES

DATE/TIME OF TEST: 12/01/2018 12:30:04

Diagnostics Check	OK	12:15	
Air Blank	0.000	12:15	
Control Test	0.080	12:15	
Air Blank	0.000	12:16	
Subject Sample	# 1 0.052	12:19	Volume Not Met
Breath Volume	0.117	12:19	
Air Blank	0.000	12:20	
Air Blank	0.000	12:22	
Subject Sample	# 2 0.048	12:24	
Breath Volume	1.320	12:24	
Air Blank	0.000	12:24	
Air Blank	0.000	12:26	
Subject Sample	# 3 0.052	12:28	
Breath Volume	1.320	12:28	
Air Blank	0.000	12:29	
Control Test	0.079	12:29	
Air Blank	0.000	12:29	
Diagnostics Check	OK	12:30	

Data Download Date/Time: 12/27/2018 13:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOSA,JONATHAN,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: SOSA,JONATHAN
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: HERRERA VILLANUEVA

DATE/TIME OF TEST: 12/03/2018 03:26:44

Diagnostics Check	OK	03:23	
Air Blank	0.000	03:24	
Control Test	0.080	03:24	
Air Blank	0.000	03:25	
Subject Sample	# 1	03:25	Subject Test Refused
Breath Volume	0.000	03:25	
Air Blank	0.000	03:25	
Control Test	0.080	03:26	
Air Blank	0.000	03:26	
Diagnostics Check	OK	03:26	

Data Download Date/Time: 12/27/2018 13:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CREWS,CURTIS,L
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: SOSA,JONATHAN
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 04:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: FORDE

DATE/TIME OF TEST: 12/03/2018 05:05:39

Diagnostics Check	OK	05:02	
Air Blank	0.000	05:02	
Control Test	0.080	05:02	
Air Blank	0.000	05:03	
Subject Sample	# 1	05:04	Subject Test Refused
Breath Volume	0.000	05:04	
Air Blank	0.000	05:04	
Control Test	0.080	05:05	
Air Blank	0.000	05:05	
Diagnostics Check	OK	05:05	

Data Download Date/Time: 12/27/2018 13:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAIGE,LEON,I
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: PAIGE,LEON
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 22:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: BUSTAMANTE

DATE/TIME OF TEST: 12/06/2018 23:08:07

Diagnostics Check	OK	22:58
Air Blank	0.000	22:58
Control Test	0.080	22:58
Air Blank	0.000	22:59
Subject Sample # 1	0.218	23:02
Breath Volume	2.023	23:02
Air Blank	0.000	23:03
Air Blank	0.000	23:05
Subject Sample # 2	0.218	23:06
Breath Volume	2.792	23:06
Air Blank	0.000	23:07
Control Test	0.076	23:07
Air Blank	0.000	23:08
Diagnostics Check	OK	23:08

Data Download Date/Time: 12/27/2018 13:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAIGE,LEON,I
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: PAIGE,LEON
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: TEJEDA RODRIGUEZ

DATE/TIME OF TEST: 12/10/2018 03:38:33

Diagnostics Check	OK	03:31
Air Blank	0.000	03:32
Control Test	0.081	03:32
Air Blank	0.000	03:33
Subject Sample # 1	0.180	03:33
Breath Volume	1.898	03:33
Air Blank	0.000	03:34
Air Blank	0.000	03:36
Subject Sample # 2	0.167	03:36
Breath Volume	2.230	03:36
Air Blank	0.000	03:37
Control Test	0.078	03:37
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

Data Download Date/Time: 12/27/2018 13:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAIGE,LEON,I
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: PAIGE,LEON
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: FIGUEROA

DATE/TIME OF TEST: 12/15/2018 04:21:08

Diagnostics Check	OK	04:11
Air Blank	0.000	04:12
Control Test	0.080	04:12
Air Blank	0.000	04:12
Subject Sample # 1	0.204	04:13
Breath Volume	2.488	04:13
Air Blank	0.000	04:14
Air Blank	0.000	04:15
Subject Sample # 2	0.182	04:16
Breath Volume	4.367	04:16
Air Blank	0.000	04:17
Air Blank	0.000	04:18
Subject Sample # 3	0.201	04:19
Breath Volume	3.578	04:19
Air Blank	0.000	04:20
Control Test	0.080	04:20
Air Blank	0.000	04:21
Diagnostics Check	OK	04:21

Data Download Date/Time: 12/27/2018 13:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAIGE,LEON,I
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: PAIGE,LEON
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 08:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:54

SUBJECT'S LAST NAME: BALL

DATE/TIME OF TEST: 12/15/2018 09:20:33

Diagnostics Check	OK	09:14
Air Blank	0.000	09:14
Control Test	0.080	09:15
Air Blank	0.000	09:15
Subject Sample # 1	0.147	09:15
Breath Volume	3.144	09:15
Air Blank	0.000	09:16
Air Blank	0.000	09:18
Subject Sample # 2	0.148	09:18
Breath Volume	3.046	09:18
Air Blank	0.000	09:19
Control Test	0.080	09:19
Air Blank	0.000	09:20
Diagnostics Check	OK	09:20

Data Download Date/Time: 12/27/2018 13:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAIGE,LEON,I
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: PAIGE,LEON
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: VELIZ TORRES

DATE/TIME OF TEST: 12/16/2018 03:50:58

Diagnostics Check	OK	03:44
Air Blank	0.000	03:44
Control Test	0.081	03:45
Air Blank	0.000	03:45
Subject Sample # 1	0.190	03:46
Breath Volume	2.402	03:46
Air Blank	0.000	03:47
Air Blank	0.000	03:48
Subject Sample # 2	0.201	03:49
Breath Volume	2.519	03:49
Air Blank	0.000	03:50
Control Test	0.080	03:50
Air Blank	0.000	03:50
Diagnostics Check	OK	03:50

Data Download Date/Time: 12/27/2018 13:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELENDEZ,DANIEL,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: RIVERA,BRIAN
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: REID

DATE/TIME OF TEST: 12/21/2018 04:07:59

Diagnostics Check	OK	03:59
Air Blank	0.000	03:59
Control Test	0.080	04:00
Air Blank	0.000	04:00
Subject Sample # 1	0.186	04:01
Breath Volume	1.875	04:01
Air Blank	0.000	04:02
Air Blank	0.000	04:04
Subject Sample # 2	0.173	04:06
Breath Volume	1.675	04:06
Air Blank	0.000	04:07
Control Test	0.079	04:07
Air Blank	0.000	04:07
Diagnostics Check	OK	04:07

Data Download Date/Time: 12/27/2018 13:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONFESSORE,STEVE,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: COSNER,STEVE
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: DVORYANINOVA

DATE/TIME OF TEST: 12/21/2018 05:24:48

Diagnostics Check	OK	05:21	
Air Blank	0.000	05:22	
Control Test	0.080	05:22	
Air Blank	0.000	05:23	
Subject Sample	# 1	05:23	Subject Test Refused
Breath Volume	0.000	05:23	
Air Blank	0.000	05:23	
Control Test	0.080	05:24	
Air Blank	0.000	05:24	
Diagnostics Check	OK	05:24	

Data Download Date/Time: 12/27/2018 13:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CREWS,CURTIS,L
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: SOSA,JONATHAN
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: DASSONVILLE

DATE/TIME OF TEST: 12/22/2018 02:42:48

Diagnostics Check	OK	02:34
Air Blank	0.000	02:34
Control Test	0.081	02:35
Air Blank	0.000	02:35
Subject Sample # 1	0.232	02:37
Breath Volume	2.300	02:37
Air Blank	0.000	02:38
Air Blank	0.000	02:39
Subject Sample # 2	0.213	02:40
Breath Volume	1.425	02:40
Air Blank	0.000	02:41
Control Test	0.075	02:42
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 12/27/2018 13:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAIGE,LEON,I
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: PAIGE,LEON
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 04:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: MIRHOSSEINI

DATE/TIME OF TEST: 12/23/2018 05:09:25

Diagnostics Check	OK	05:02
Air Blank	0.000	05:03
Control Test	0.082	05:03
Air Blank	0.000	05:04
Subject Sample # 1	0.162	05:04
Breath Volume	1.433	05:04
Air Blank	0.000	05:05
Air Blank	0.000	05:07
Subject Sample # 2	0.159	05:07
Breath Volume	2.121	05:07
Air Blank	0.000	05:08
Control Test	0.081	05:08
Air Blank	0.000	05:09
Diagnostics Check	OK	05:09

Data Download Date/Time: 12/27/2018 13:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAIGE,LEON,I
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: PAIGE,LEON
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 07:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:42

SUBJECT'S LAST NAME: SOMARRIBA PINEDA

DATE/TIME OF TEST: 12/23/2018 08:09:39

Diagnostics Check	OK	08:03
Air Blank	0.000	08:03
Control Test	0.081	08:03
Air Blank	0.000	08:04
Subject Sample # 1	0.158	08:05
Breath Volume	1.507	08:05
Air Blank	0.000	08:05
Air Blank	0.000	08:07
Subject Sample # 2	0.149	08:07
Breath Volume	1.351	08:07
Air Blank	0.000	08:08
Control Test	0.081	08:09
Air Blank	0.000	08:09
Diagnostics Check	OK	08:09

Data Download Date/Time: 12/27/2018 13:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUMPHRY,KERRY,R
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: ANDERSON,DONALD
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 08:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:31

SUBJECT'S LAST NAME: GOLDES

DATE/TIME OF TEST: 12/26/2018 09:00:33

Diagnostics Check	OK	08:53
Air Blank	0.000	08:54
Control Test	0.081	08:54
Air Blank	0.000	08:55
Subject Sample # 1	0.324	08:55
Breath Volume	1.816	08:55
Air Blank	0.000	08:56
Air Blank	0.000	08:58
Subject Sample # 2	0.316	08:58
Breath Volume	3.000	08:58
Air Blank	0.000	08:59
Control Test	0.080	08:59
Air Blank	0.000	09:00
Diagnostics Check	OK	09:00

Data Download Date/Time: 12/27/2018 13:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN,ANTHONY,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: NONE,NONE
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: PROB/PAROL
ARREST TIME: 07:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:20

SUBJECT'S LAST NAME: OTOOLE

DATE/TIME OF TEST: 12/27/2018 07:46:30

Diagnostics Check	OK	07:40
Air Blank	0.000	07:40
Control Test	0.081	07:41
Air Blank	0.000	07:41
Subject Sample # 1	0.000	07:42
Breath Volume	3.074	07:42
Air Blank	0.000	07:42
Air Blank	0.000	07:44
Subject Sample # 2	0.000	07:45
Breath Volume	2.757	07:45
Air Blank	0.000	07:45
Control Test	0.081	07:45
Air Blank	0.000	07:46
Diagnostics Check	OK	07:46

Data Download Date/Time: 12/27/2018 13:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,BARRY,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: NIXON,BARRY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 11/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 12/05/2018 22:26:55

Diagnostics Check	OK	22:20
Air Blank	0.000	22:21
Control Test	0.082	22:21
Air Blank	0.000	22:22
Subject Sample # 1	0.202	22:22
Breath Volume	2.687	22:22
Air Blank	0.000	22:23
Air Blank	0.000	22:25
Subject Sample # 2	0.189	22:25
Breath Volume	2.402	22:25
Air Blank	0.000	22:26
Control Test	0.081	22:26
Air Blank	0.000	22:26
Diagnostics Check	OK	22:26

Data Download Date/Time: 12/19/2018 18:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,BARRY,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: NIXON,BARRY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 11/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: MATHEW

DATE/TIME OF TEST: 12/06/2018 17:58:35

Diagnostics Check	OK	17:51
Air Blank	0.000	17:52
Control Test	0.081	17:52
Air Blank	0.000	17:53
Subject Sample # 1	0.293	17:54
Breath Volume	1.875	17:54
Air Blank	0.000	17:54
Air Blank	0.000	17:56
Subject Sample # 2	0.301	17:57
Breath Volume	2.007	17:57
Air Blank	0.000	17:57
Control Test	0.081	17:57
Air Blank	0.000	17:58
Diagnostics Check	OK	17:58

Data Download Date/Time: 12/19/2018 18:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,BARRY,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: NIXON,BARRY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 11/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: KOTT

DATE/TIME OF TEST: 12/09/2018 02:48:34

Diagnostics Check	OK	02:47	
Air Blank		02:48	Ambient Fail
Air Blank	0.000	02:48	

Data Download Date/Time: 12/19/2018 18:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,BARRY,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: NIXON,BARRY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 11/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: KOTT

DATE/TIME OF TEST: 12/09/2018 02:50:53

Diagnostics Check	OK	02:49	
Air Blank		02:50	Ambient Fail
Air Blank	0.000	02:50	

Data Download Date/Time: 12/19/2018 18:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,BARRY,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: NIXON,BARRY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 11/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: KOTT

DATE/TIME OF TEST: 12/09/2018 02:53:09

Diagnostics Check	OK	02:52	
Air Blank		02:52	Ambient Fail
Air Blank	0.000	02:53	

Data Download Date/Time: 12/19/2018 18:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,BARRY,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: HAMILTON,BRANDON
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 12/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: LOINES

DATE/TIME OF TEST: 12/23/2018 01:21:55

Diagnostics Check	OK	01:15
Air Blank	0.000	01:16
Control Test	0.081	01:16
Air Blank	0.000	01:16
Subject Sample # 1	0.129	01:17
Breath Volume	1.988	01:17
Air Blank	0.000	01:17
Air Blank	0.000	01:19
Subject Sample # 2	0.126	01:20
Breath Volume	2.769	01:20
Air Blank	0.000	01:20
Control Test	0.082	01:21
Air Blank	0.000	01:21
Diagnostics Check	OK	01:21

Data Download Date/Time: 01/09/2019 20:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MCLAUGHLIN, W
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: BOWDEN

DATE/TIME OF TEST: 12/18/2018 20:22:23

Diagnostics Check	OK	20:16
Air Blank	0.000	20:16
Control Test	0.079	20:16
Air Blank	0.000	20:17
Subject Sample # 1	0.154	20:17
Breath Volume	1.820	20:17
Air Blank	0.000	20:18
Air Blank	0.000	20:20
Subject Sample # 2	0.157	20:20
Breath Volume	1.867	20:20
Air Blank	0.000	20:21
Control Test	0.079	20:21
Air Blank	0.000	20:22
Diagnostics Check	OK	20:22

Data Download Date/Time: 01/10/2019 17:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MACIAS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:02

SUBJECT'S LAST NAME: KETCHAM

DATE/TIME OF TEST: 12/18/2018 21:29:42

Diagnostics Check	OK	21:22
Air Blank	0.000	21:23
Control Test	0.079	21:23
Air Blank	0.000	21:24
Subject Sample # 1	0.000	21:25
Breath Volume	3.136	21:25
Air Blank	0.000	21:25
Air Blank	0.000	21:27
Subject Sample # 2	0.000	21:28
Breath Volume	3.382	21:28
Air Blank	0.000	21:28
Control Test	0.079	21:29
Air Blank	0.000	21:29
Diagnostics Check	OK	21:29

Data Download Date/Time: 01/10/2019 17:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: YEPEZ

DATE/TIME OF TEST: 12/19/2018 03:03:29

Diagnostics Check	OK	02:56
Air Blank	0.000	02:56
Control Test	0.079	02:57
Air Blank	0.000	02:57
Subject Sample # 1	0.126	02:59
Breath Volume	1.546	02:59
Air Blank	0.000	02:59
Air Blank	0.000	03:01
Subject Sample # 2	0.125	03:01
Breath Volume	1.578	03:01
Air Blank	0.000	03:02
Control Test	0.079	03:02
Air Blank	0.000	03:03
Diagnostics Check	OK	03:03

Data Download Date/Time: 01/10/2019 17:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKIPPER,ROGER,G
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: REICHERT,C 4703
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 13:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:04

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 12/19/2018 14:26:26

Diagnostics Check	OK	14:24	
Air Blank	0.000	14:25	
Control Test	0.079	14:25	
Air Blank		14:25	Sequence Aborted
Air Blank	0.000	14:26	

Data Download Date/Time: 01/10/2019 17:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKIPPER,ROGER,G
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: REICHERT,C 4703
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 13:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:27

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 12/19/2018 14:55:27

Diagnostics Check	OK	14:47
Air Blank	0.000	14:48
Control Test	0.078	14:48
Air Blank	0.000	14:49
Subject Sample # 1	0.290	14:50
Breath Volume	2.312	14:50
Air Blank	0.000	14:51
Air Blank	0.000	14:52
Subject Sample # 2	0.308	14:53
Breath Volume	2.753	14:53
Air Blank	0.000	14:54
Control Test	0.078	14:54
Air Blank	0.000	14:55
Diagnostics Check	OK	14:55

Data Download Date/Time: 01/10/2019 17:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: POVEY

DATE/TIME OF TEST: 12/19/2018 20:28:35

Diagnostics Check	OK	20:25	
Air Blank	0.000	20:25	
Control Test	0.078	20:25	
Air Blank	0.000	20:26	
Subject Sample	# 1	20:27	Subject Test Refused
Breath Volume	0.000	20:27	
Air Blank	0.000	20:27	
Control Test	0.079	20:27	
Air Blank	0.000	20:28	
Diagnostics Check	OK	20:28	

Data Download Date/Time: 01/10/2019 17:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: STALKER

DATE/TIME OF TEST: 12/21/2018 03:54:20

Diagnostics Check	OK	03:47
Air Blank	0.000	03:47
Control Test	0.079	03:48
Air Blank	0.000	03:48
Subject Sample # 1	0.096	03:49
Breath Volume	2.105	03:49
Air Blank	0.000	03:49
Air Blank	0.000	03:51
Subject Sample # 2	0.099	03:52
Breath Volume	3.367	03:52
Air Blank	0.000	03:53
Control Test	0.079	03:53
Air Blank	0.000	03:54
Diagnostics Check	OK	03:54

Data Download Date/Time: 01/10/2019 17:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKIPPER,ROGER,G
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: PAYNTER,H 4303
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 05:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:12

SUBJECT'S LAST NAME: GINORIO RIOS

DATE/TIME OF TEST: 12/21/2018 07:39:20

Diagnostics Check	OK	07:32
Air Blank	0.000	07:33
Control Test	0.078	07:33
Air Blank	0.000	07:34
Subject Sample # 1	0.216	07:34
Breath Volume	1.617	07:34
Air Blank	0.000	07:35
Air Blank	0.000	07:36
Subject Sample # 2	0.214	07:37
Breath Volume	1.570	07:37
Air Blank	0.000	07:38
Control Test	0.078	07:38
Air Blank	0.000	07:39
Diagnostics Check	OK	07:39

Data Download Date/Time: 01/10/2019 17:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DE JESUS,V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: DOMINGO-JIMENEZ

DATE/TIME OF TEST: 12/21/2018 23:41:27

Diagnostics Check	OK	23:33
Air Blank	0.000	23:33
Control Test	0.079	23:34
Air Blank	0.000	23:34
Subject Sample # 1	0.137	23:36
Breath Volume	2.167	23:36
Air Blank	0.000	23:36
Air Blank	0.000	23:38
Subject Sample # 2	0.132	23:39
Breath Volume	1.917	23:39
Air Blank	0.000	23:40
Control Test	0.078	23:40
Air Blank	0.000	23:41
Diagnostics Check	OK	23:41

Data Download Date/Time: 01/10/2019 17:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES AVERY, DAVARIA, C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SCHILLER, E
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 12/22/2018 02:37:36

Diagnostics Check	OK	02:31
Air Blank	0.000	02:31
Control Test	0.079	02:32
Air Blank	0.000	02:32
Subject Sample # 1	0.129	02:33
Breath Volume	2.078	02:33
Air Blank	0.000	02:33
Air Blank	0.000	02:35
Subject Sample # 2	0.135	02:36
Breath Volume	2.183	02:36
Air Blank	0.000	02:36
Control Test	0.078	02:37
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

Data Download Date/Time: 01/10/2019 17:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: REID,W
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: SANTIAGO CIENA

DATE/TIME OF TEST: 12/22/2018 04:32:00

Diagnostics Check	OK	04:28	
Air Blank	0.000	04:29	
Control Test	0.079	04:29	
Air Blank	0.000	04:29	
Subject Sample	# 1	04:30	Subject Test Refused
Breath Volume	0.000	04:30	
Air Blank	0.000	04:31	
Control Test	0.079	04:31	
Air Blank	0.000	04:31	
Diagnostics Check	OK	04:32	

Data Download Date/Time: 01/10/2019 17:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HOWARD,D
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: BIGAS-SOTO

DATE/TIME OF TEST: 12/22/2018 06:07:49

Diagnostics Check	OK	06:01
Air Blank	0.000	06:02
Control Test	0.079	06:02
Air Blank	0.000	06:02
Subject Sample # 1	0.169	06:03
Breath Volume	1.339	06:03
Air Blank	0.000	06:03
Air Blank	0.000	06:05
Subject Sample # 2	0.175	06:06
Breath Volume	2.218	06:06
Air Blank	0.000	06:06
Control Test	0.078	06:07
Air Blank	0.000	06:07
Diagnostics Check	OK	06:07

Data Download Date/Time: 01/10/2019 17:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MACIAS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:01

SUBJECT'S LAST NAME: DINWOODIE

DATE/TIME OF TEST: 12/22/2018 23:29:23

Diagnostics Check	OK	23:22
Air Blank	0.000	23:22
Control Test	0.080	23:22
Air Blank	0.000	23:23
Subject Sample # 1	0.279	23:24
Breath Volume	2.402	23:24
Air Blank	0.000	23:25
Air Blank	0.000	23:27
Subject Sample # 2	0.275	23:27
Breath Volume	1.144	23:27
Air Blank	0.000	23:28
Control Test	0.078	23:28
Air Blank	0.000	23:29
Diagnostics Check	OK	23:29

Data Download Date/Time: 01/10/2019 17:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIMS,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 12/23/2018 04:44:09

Diagnostics Check	OK	04:37
Air Blank	0.000	04:38
Control Test	0.080	04:38
Air Blank	0.000	04:38
Subject Sample # 1	0.195	04:39
Breath Volume	3.085	04:39
Air Blank	0.000	04:40
Air Blank	0.000	04:41
Subject Sample # 2	0.199	04:42
Breath Volume	3.062	04:42
Air Blank	0.000	04:43
Control Test	0.079	04:43
Air Blank	0.000	04:44
Diagnostics Check	OK	04:44

Data Download Date/Time: 01/10/2019 17:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,BARBARA,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DE JESUS,V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 07:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:00

SUBJECT'S LAST NAME: VILIERE

DATE/TIME OF TEST: 12/23/2018 08:26:49

Diagnostics Check	OK	08:20
Air Blank	0.000	08:20
Control Test	0.079	08:20
Air Blank	0.000	08:21
Subject Sample # 1	0.174	08:21
Breath Volume	2.765	08:21
Air Blank	0.000	08:22
Air Blank	0.000	08:24
Subject Sample # 2	0.186	08:25
Breath Volume	2.433	08:25
Air Blank	0.000	08:25
Control Test	0.079	08:26
Air Blank	0.000	08:26
Diagnostics Check	OK	08:26

Data Download Date/Time: 01/10/2019 17:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 12/23/2018 22:29:42

Diagnostics Check	OK	22:22
Air Blank	0.000	22:22
Control Test	0.079	22:23
Air Blank	0.000	22:23
Subject Sample # 1	0.120	22:25
Breath Volume	2.894	22:25
Air Blank	0.000	22:25
Air Blank	0.000	22:27
Subject Sample # 2	0.121	22:28
Breath Volume	2.996	22:28
Air Blank	0.000	22:28
Control Test	0.079	22:29
Air Blank	0.000	22:29
Diagnostics Check	OK	22:29

Data Download Date/Time: 01/10/2019 17:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 12/23/2018 23:57:26

Diagnostics Check	OK	23:50
Air Blank	0.000	23:50
Control Test	0.078	23:51
Air Blank	0.000	23:51
Subject Sample # 1	0.212	23:52
Breath Volume	2.777	23:52
Air Blank	0.000	23:53
Air Blank	0.000	23:55
Subject Sample # 2	0.216	23:55
Breath Volume	2.105	23:55
Air Blank	0.000	23:56
Control Test	0.079	23:56
Air Blank	0.000	23:57
Diagnostics Check	OK	23:57

Data Download Date/Time: 01/10/2019 17:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SCHAEFER,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:36

SUBJECT'S LAST NAME: SUAREZ ORTIZ

DATE/TIME OF TEST: 12/24/2018 20:03:09

Diagnostics Check	OK	19:56
Air Blank	0.000	19:57
Control Test	0.079	19:57
Air Blank	0.000	19:57
Subject Sample # 1	0.239	19:58
Breath Volume	3.000	19:58
Air Blank	0.000	19:59
Air Blank	0.000	20:01
Subject Sample # 2	0.242	20:01
Breath Volume	2.433	20:01
Air Blank	0.000	20:02
Control Test	0.079	20:02
Air Blank	0.000	20:03
Diagnostics Check	OK	20:03

Data Download Date/Time: 01/10/2019 17:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PECYLAK,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: CULBERSON

DATE/TIME OF TEST: 12/24/2018 23:16:14

Diagnostics Check	OK	23:09
Air Blank	0.000	23:09
Control Test	0.080	23:10
Air Blank	0.000	23:10
Subject Sample # 1	0.082	23:11
Breath Volume	1.351	23:11
Air Blank	0.000	23:12
Air Blank	0.000	23:13
Subject Sample # 2	0.082	23:14
Breath Volume	1.375	23:14
Air Blank	0.000	23:15
Control Test	0.080	23:15
Air Blank	0.000	23:16
Diagnostics Check	OK	23:16

Data Download Date/Time: 01/10/2019 17:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,REBECCA,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BOOKER,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:41

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 12/25/2018 16:08:37

Diagnostics Check	OK	16:01
Air Blank	0.000	16:02
Control Test	0.079	16:02
Air Blank	0.000	16:03
Subject Sample # 1	0.241	16:03
Breath Volume	1.636	16:03
Air Blank	0.000	16:04
Air Blank	0.000	16:06
Subject Sample # 2	0.233	16:07
Breath Volume	1.835	16:07
Air Blank	0.000	16:07
Control Test	0.078	16:08
Air Blank	0.000	16:08
Diagnostics Check	OK	16:08

Data Download Date/Time: 01/10/2019 17:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKIPPER,ROGER,G
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DELACRUZ,A 251318
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 09:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:09

SUBJECT'S LAST NAME: PERRY

DATE/TIME OF TEST: 12/26/2018 10:39:02

Diagnostics Check	OK	10:31
Air Blank	0.000	10:31
Control Test	0.079	10:32
Air Blank	0.000	10:32
Subject Sample # 1	0.285	10:34
Breath Volume	1.582	10:34
Air Blank	0.000	10:35
Air Blank	0.000	10:36
Subject Sample # 2	0.282	10:37
Breath Volume	1.941	10:37
Air Blank	0.000	10:38
Control Test	0.077	10:38
Air Blank	0.000	10:38
Diagnostics Check	OK	10:39

Data Download Date/Time: 01/10/2019 17:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:56

SUBJECT'S LAST NAME: JAMES

DATE/TIME OF TEST: 12/27/2018 20:20:38

Diagnostics Check	OK	20:16	
Air Blank	0.000	20:17	
Control Test	0.079	20:17	
Air Blank	0.000	20:18	
Subject Sample	# 1	20:19	Interferent Detect
Breath Volume	1.632	20:19	
Air Blank	0.000	20:19	
Control Test	0.079	20:20	
Air Blank	0.000	20:20	
Diagnostics Check	OK	20:20	

Data Download Date/Time: 01/10/2019 17:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FLORES,C
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: DURHAM

DATE/TIME OF TEST: 12/27/2018 23:39:39

Diagnostics Check	OK	23:33
Air Blank	0.000	23:33
Control Test	0.080	23:33
Air Blank	0.000	23:34
Subject Sample # 1	0.000	23:35
Breath Volume	1.812	23:35
Air Blank	0.000	23:35
Air Blank	0.000	23:37
Subject Sample # 2	0.000	23:38
Breath Volume	1.101	23:38
Air Blank	0.000	23:38
Control Test	0.079	23:39
Air Blank	0.000	23:39
Diagnostics Check	OK	23:39

Data Download Date/Time: 01/10/2019 17:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: MELGAR

DATE/TIME OF TEST: 12/28/2018 21:58:53

Diagnostics Check	OK	21:51
Air Blank	0.000	21:51
Control Test	0.079	21:52
Air Blank	0.000	21:52
Subject Sample # 1	0.075	21:54
Breath Volume	1.496	21:54
Air Blank	0.000	21:54
Air Blank	0.000	21:56
Subject Sample # 2	0.073	21:57
Breath Volume	1.558	21:57
Air Blank	0.000	21:57
Control Test	0.079	21:58
Air Blank	0.000	21:58
Diagnostics Check	OK	21:58

Data Download Date/Time: 01/10/2019 17:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,REBECCA,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JOHNSON,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 06:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:27

SUBJECT'S LAST NAME: MILLS

DATE/TIME OF TEST: 12/29/2018 08:03:36

Diagnostics Check	OK	07:48	
Air Blank	0.000	07:48	
Control Test	0.078	07:48	
Air Blank	0.000	07:49	
Subject Sample	# 1 0.130	07:52	Volume Not Met
Breath Volume	0.468	07:52	
Air Blank	0.000	07:53	
Air Blank	0.000	07:55	
Subject Sample	# 2 0.142	07:58	
Breath Volume	1.175	07:58	
Air Blank	0.000	07:58	
Air Blank	0.000	08:00	
Subject Sample	# 3 0.139	08:02	
Breath Volume	1.468	08:02	
Air Blank	0.000	08:02	
Control Test	0.077	08:02	
Air Blank	0.000	08:03	
Diagnostics Check	OK	08:03	

Data Download Date/Time: 01/10/2019 17:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: COSTA,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: SCAGLIONE

DATE/TIME OF TEST: 12/29/2018 20:06:08

Diagnostics Check	OK	20:00
Air Blank	0.000	20:00
Control Test	0.079	20:01
Air Blank	0.000	20:01
Subject Sample # 1	0.000	20:01
Breath Volume	2.179	20:01
Air Blank	0.000	20:02
Air Blank	0.000	20:04
Subject Sample # 2	0.000	20:04
Breath Volume	1.640	20:04
Air Blank	0.000	20:05
Control Test	0.079	20:05
Air Blank	0.000	20:06
Diagnostics Check	OK	20:06

Data Download Date/Time: 01/10/2019 17:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DRISCOLL,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: WISE

DATE/TIME OF TEST: 12/29/2018 23:14:23

Diagnostics Check	OK	23:07
Air Blank	0.000	23:07
Control Test	0.079	23:08
Air Blank	0.000	23:08
Subject Sample # 1	0.040	23:09
Breath Volume	2.058	23:09
Air Blank	0.000	23:10
Air Blank	0.000	23:12
Subject Sample # 2	0.041	23:12
Breath Volume	2.304	23:12
Air Blank	0.000	23:13
Control Test	0.080	23:13
Air Blank	0.000	23:14
Diagnostics Check	OK	23:14

Data Download Date/Time: 01/10/2019 17:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES AVERY, DAVARIA, C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER, A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: TOWNSEND

DATE/TIME OF TEST: 12/31/2018 00:49:07

Diagnostics Check	OK	00:42
Air Blank	0.000	00:43
Control Test	0.079	00:43
Air Blank	0.000	00:44
Subject Sample # 1	0.147	00:44
Breath Volume	2.070	00:44
Air Blank	0.000	00:45
Air Blank	0.000	00:47
Subject Sample # 2	0.142	00:47
Breath Volume	1.542	00:47
Air Blank	0.000	00:48
Control Test	0.078	00:48
Air Blank	0.000	00:49
Diagnostics Check	OK	00:49

Data Download Date/Time: 01/10/2019 17:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ARCHAMBEAU,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: MAKKAR

DATE/TIME OF TEST: 12/31/2018 03:29:37

Diagnostics Check	OK	03:22
Air Blank	0.000	03:23
Control Test	0.079	03:23
Air Blank	0.000	03:24
Subject Sample # 1	0.243	03:24
Breath Volume	3.886	03:24
Air Blank	0.000	03:25
Air Blank	0.000	03:27
Subject Sample # 2	0.239	03:27
Breath Volume	3.753	03:27
Air Blank	0.000	03:28
Control Test	0.079	03:29
Air Blank	0.000	03:29
Diagnostics Check	OK	03:29

Data Download Date/Time: 01/10/2019 17:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,REBECCA,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HIV,C
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 07:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:48

SUBJECT'S LAST NAME: ADEKANMBI

DATE/TIME OF TEST: 12/31/2018 09:17:48

Diagnostics Check	OK	09:09
Air Blank	0.000	09:10
Control Test	0.078	09:10
Air Blank	0.000	09:11
Subject Sample # 1	0.045	09:13
Breath Volume	2.625	09:13
Air Blank	0.000	09:13
Air Blank	0.000	09:15
Subject Sample # 2	0.043	09:16
Breath Volume	2.164	09:16
Air Blank	0.000	09:16
Control Test	0.078	09:17
Air Blank	0.000	09:17
Diagnostics Check	OK	09:17

Data Download Date/Time: 01/10/2019 17:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FLORES,C
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: SKYERS

DATE/TIME OF TEST: 12/19/2018 01:35:24

Diagnostics Check	OK	01:28
Air Blank	0.000	01:28
Control Test	0.081	01:29
Air Blank	0.000	01:29
Subject Sample # 1	0.000	01:31
Breath Volume	2.492	01:31
Air Blank	0.000	01:31
Air Blank	0.000	01:33
Subject Sample # 2	0.000	01:34
Breath Volume	1.863	01:34
Air Blank	0.000	01:34
Control Test	0.081	01:34
Air Blank	0.000	01:35
Diagnostics Check	OK	01:35

Data Download Date/Time: 01/10/2019 19:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: MANHIN

DATE/TIME OF TEST: 12/20/2018 01:42:30

Diagnostics Check	OK	01:35
Air Blank	0.000	01:36
Control Test	0.081	01:36
Air Blank	0.000	01:37
Subject Sample # 1	0.000	01:37
Breath Volume	2.945	01:37
Air Blank	0.000	01:38
Air Blank	0.000	01:40
Subject Sample # 2	0.000	01:41
Breath Volume	3.460	01:41
Air Blank	0.000	01:41
Control Test	0.080	01:41
Air Blank	0.000	01:42
Diagnostics Check	OK	01:42

Data Download Date/Time: 01/10/2019 19:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SCHAEFER,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: BORGOS-COLON

DATE/TIME OF TEST: 12/20/2018 04:53:59

Diagnostics Check	OK	04:47
Air Blank	0.000	04:47
Control Test	0.079	04:47
Air Blank	0.000	04:48
Subject Sample # 1	0.139	04:49
Breath Volume	2.277	04:49
Air Blank	0.000	04:50
Air Blank	0.000	04:51
Subject Sample # 2	0.145	04:52
Breath Volume	1.898	04:52
Air Blank	0.000	04:53
Control Test	0.079	04:53
Air Blank	0.000	04:53
Diagnostics Check	OK	04:53

Data Download Date/Time: 01/10/2019 19:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DRISCOLL,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:37

SUBJECT'S LAST NAME: GEIER

DATE/TIME OF TEST: 12/20/2018 22:04:32

Diagnostics Check	OK	21:57
Air Blank	0.000	21:58
Control Test	0.081	21:58
Air Blank	0.000	21:59
Subject Sample # 1	0.167	21:59
Breath Volume	2.632	21:59
Air Blank	0.000	22:00
Air Blank	0.000	22:02
Subject Sample # 2	0.174	22:02
Breath Volume	2.847	22:02
Air Blank	0.000	22:03
Control Test	0.079	22:03
Air Blank	0.000	22:04
Diagnostics Check	OK	22:04

Data Download Date/Time: 01/10/2019 19:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES AVERY, DAVARIA, C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: COSTA, A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 12/21/2018 05:11:30

Diagnostics Check	OK	05:05
Air Blank	0.000	05:05
Control Test	0.080	05:06
Air Blank	0.000	05:06
Subject Sample # 1	0.160	05:07
Breath Volume	2.289	05:07
Air Blank	0.000	05:07
Air Blank	0.000	05:09
Subject Sample # 2	0.156	05:09
Breath Volume	2.328	05:09
Air Blank	0.000	05:10
Control Test	0.079	05:10
Air Blank	0.000	05:11
Diagnostics Check	OK	05:11

Data Download Date/Time: 01/10/2019 19:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MOREJON,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 12/21/2018 23:32:32

Diagnostics Check	OK	23:17	
Air Blank	0.000	23:17	
Control Test	0.082	23:17	
Air Blank	0.000	23:18	
Subject Sample	# 1 0.194	23:19	
Breath Volume	2.140	23:19	
Air Blank	0.000	23:20	
Air Blank	0.000	23:22	
Subject Sample	# 2	23:25	No Sample Provided
Breath Volume	0.000	23:25	
Air Blank	0.000	23:26	
Air Blank	0.000	23:27	
Subject Sample	# 3 0.000	23:30	Volume Not Met
Breath Volume	0.015	23:30	
Air Blank	0.000	23:31	
Control Test	0.079	23:31	
Air Blank	0.000	23:32	
Diagnostics Check	OK	23:32	

Data Download Date/Time: 01/10/2019 19:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MOREJON,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 12/21/2018 23:47:10

Diagnostics Check	OK	23:35	
Air Blank	0.000	23:36	
Control Test	0.079	23:36	
Air Blank	0.000	23:37	
Subject Sample	# 1	23:40	No Sample Provided
Breath Volume	0.000	23:40	
Air Blank	0.000	23:40	
Air Blank	0.000	23:42	
Subject Sample	# 2	23:45	No Sample Provided
Breath Volume	0.000	23:45	
Air Blank	0.000	23:46	
Control Test	0.080	23:46	
Air Blank	0.000	23:47	
Diagnostics Check	OK	23:47	

Data Download Date/Time: 01/10/2019 19:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES AVERY, DAVARIA, C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 12/22/2018 01:17:12

Diagnostics Check	OK	01:10
Air Blank	0.000	01:11
Control Test	0.081	01:11
Air Blank	0.000	01:11
Subject Sample # 1	0.080	01:12
Breath Volume	1.511	01:12
Air Blank	0.000	01:13
Air Blank	0.000	01:15
Subject Sample # 2	0.080	01:15
Breath Volume	1.964	01:15
Air Blank	0.000	01:16
Control Test	0.080	01:16
Air Blank	0.000	01:17
Diagnostics Check	OK	01:17

Data Download Date/Time: 01/10/2019 19:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BODMER,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: CANDELARIO

DATE/TIME OF TEST: 12/22/2018 05:11:27

Diagnostics Check	OK	05:04
Air Blank	0.000	05:05
Control Test	0.080	05:05
Air Blank	0.000	05:06
Subject Sample # 1	0.209	05:07
Breath Volume	1.621	05:07
Air Blank	0.000	05:07
Air Blank	0.000	05:09
Subject Sample # 2	0.199	05:09
Breath Volume	1.617	05:09
Air Blank	0.000	05:10
Control Test	0.081	05:10
Air Blank	0.000	05:11
Diagnostics Check	OK	05:11

Data Download Date/Time: 01/10/2019 19:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES AVERY, DAVARIA, C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRIZENDINE, J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:36

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 12/22/2018 06:02:44

Diagnostics Check	OK	05:56
Air Blank	0.000	05:56
Control Test	0.081	05:57
Air Blank	0.000	05:57
Subject Sample # 1	0.132	05:58
Breath Volume	2.035	05:58
Air Blank	0.000	05:58
Air Blank	0.000	06:00
Subject Sample # 2	0.134	06:01
Breath Volume	2.046	06:01
Air Blank	0.000	06:01
Control Test	0.080	06:02
Air Blank	0.000	06:02
Diagnostics Check	OK	06:02

Data Download Date/Time: 01/10/2019 19:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: NAJERA

DATE/TIME OF TEST: 12/23/2018 02:44:57

Diagnostics Check	OK	02:37
Air Blank	0.000	02:37
Control Test	0.082	02:38
Air Blank	0.000	02:38
Subject Sample # 1	0.207	02:40
Breath Volume	3.292	02:40
Air Blank	0.000	02:40
Air Blank	0.000	02:42
Subject Sample # 2	0.214	02:43
Breath Volume	3.035	02:43
Air Blank	0.000	02:44
Control Test	0.080	02:44
Air Blank	0.000	02:44
Diagnostics Check	OK	02:44

Data Download Date/Time: 01/10/2019 19:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: WORTHINGTON

DATE/TIME OF TEST: 12/23/2018 04:24:47

Diagnostics Check	OK	04:16
Air Blank	0.000	04:16
Control Test	0.082	04:17
Air Blank	0.000	04:17
Subject Sample # 1	0.137	04:18
Breath Volume	3.019	04:18
Air Blank	0.000	04:19
Air Blank	0.000	04:21
Subject Sample # 2	0.141	04:23
Breath Volume	2.449	04:23
Air Blank	0.000	04:23
Control Test	0.081	04:24
Air Blank	0.000	04:24
Diagnostics Check	OK	04:24

Data Download Date/Time: 01/10/2019 19:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: MURRAY

DATE/TIME OF TEST: 12/23/2018 04:54:02

Diagnostics Check	OK	04:44
Air Blank	0.000	04:45
Control Test	0.081	04:45
Air Blank	0.000	04:46
Subject Sample # 1	0.133	04:48
Breath Volume	1.542	04:48
Air Blank	0.000	04:48
Air Blank	0.000	04:50
Subject Sample # 2	0.136	04:52
Breath Volume	1.972	04:52
Air Blank	0.000	04:53
Control Test	0.080	04:53
Air Blank	0.000	04:53
Diagnostics Check	OK	04:54

Data Download Date/Time: 01/10/2019 19:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DE JESUS, V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:26

SUBJECT'S LAST NAME: GONZALEZ MOCTEZUMA

DATE/TIME OF TEST: 12/23/2018 20:53:08

Diagnostics Check	OK	20:46
Air Blank	0.000	20:47
Control Test	0.082	20:47
Air Blank	0.000	20:48
Subject Sample # 1	0.167	20:48
Breath Volume	2.421	20:48
Air Blank	0.000	20:49
Air Blank	0.000	20:51
Subject Sample # 2	0.166	20:51
Breath Volume	1.542	20:51
Air Blank	0.000	20:52
Control Test	0.081	20:52
Air Blank	0.000	20:53
Diagnostics Check	OK	20:53

Data Download Date/Time: 01/10/2019 19:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 12/24/2018 01:17:56

Diagnostics Check	OK	01:11
Air Blank	0.000	01:11
Control Test	0.082	01:11
Air Blank	0.000	01:12
Subject Sample # 1	0.167	01:13
Breath Volume	1.367	01:13
Air Blank	0.000	01:13
Air Blank	0.000	01:15
Subject Sample # 2	0.168	01:16
Breath Volume	1.542	01:16
Air Blank	0.000	01:17
Control Test	0.079	01:17
Air Blank	0.000	01:17
Diagnostics Check	OK	01:17

Data Download Date/Time: 01/10/2019 19:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MACIAS,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: SOLOMON

DATE/TIME OF TEST: 12/24/2018 02:09:28

Diagnostics Check	OK	02:02
Air Blank	0.000	02:03
Control Test	0.081	02:03
Air Blank	0.000	02:03
Subject Sample # 1	0.178	02:05
Breath Volume	1.777	02:05
Air Blank	0.000	02:05
Air Blank	0.000	02:07
Subject Sample # 2	0.183	02:07
Breath Volume	1.878	02:07
Air Blank	0.000	02:08
Control Test	0.080	02:08
Air Blank	0.000	02:09
Diagnostics Check	OK	02:09

Data Download Date/Time: 01/10/2019 19:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,TIMOTHY,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DE LIMA,ALEX
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:28

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 12/24/2018 15:55:56

Diagnostics Check	OK	15:49
Air Blank	0.000	15:50
Control Test	0.081	15:50
Air Blank	0.000	15:50
Subject Sample # 1	0.000	15:51
Breath Volume	2.667	15:51
Air Blank	0.000	15:51
Air Blank	0.000	15:53
Subject Sample # 2	0.000	15:54
Breath Volume	2.820	15:54
Air Blank	0.000	15:55
Control Test	0.082	15:55
Air Blank	0.000	15:55
Diagnostics Check	OK	15:55

Data Download Date/Time: 01/10/2019 19:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,BARBARA,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARRELL,Q
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:39

SUBJECT'S LAST NAME: GARCIA HERNANDEZ

DATE/TIME OF TEST: 12/24/2018 16:06:29

Diagnostics Check	OK	15:59
Air Blank	0.000	16:00
Control Test	0.080	16:00
Air Blank	0.000	16:00
Subject Sample # 1	0.338	16:01
Breath Volume	1.738	16:01
Air Blank	0.000	16:02
Air Blank	0.000	16:04
Subject Sample # 2	0.329	16:04
Breath Volume	1.769	16:04
Air Blank	0.000	16:05
Control Test	0.079	16:05
Air Blank	0.000	16:06
Diagnostics Check	OK	16:06

Data Download Date/Time: 01/10/2019 19:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: HERNANDEZ HERNANDEZ

DATE/TIME OF TEST: 12/25/2018 04:49:33

Diagnostics Check	OK	04:40
Air Blank	0.000	04:40
Control Test	0.081	04:41
Air Blank	0.000	04:41
Subject Sample # 1	0.116	04:44
Breath Volume	1.535	04:44
Air Blank	0.000	04:45
Air Blank	0.000	04:46
Subject Sample # 2	0.113	04:48
Breath Volume	1.531	04:48
Air Blank	0.000	04:48
Control Test	0.079	04:48
Air Blank	0.000	04:49
Diagnostics Check	OK	04:49

Data Download Date/Time: 01/10/2019 19:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIMPKINS,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: BERNAL AVILA

DATE/TIME OF TEST: 12/25/2018 22:48:50

Diagnostics Check	OK	22:42
Air Blank	0.000	22:42
Control Test	0.081	22:43
Air Blank	0.000	22:43
Subject Sample # 1	0.124	22:44
Breath Volume	3.531	22:44
Air Blank	0.000	22:44
Air Blank	0.000	22:46
Subject Sample # 2	0.127	22:47
Breath Volume	2.894	22:47
Air Blank	0.000	22:47
Control Test	0.080	22:48
Air Blank	0.000	22:48
Diagnostics Check	OK	22:48

Data Download Date/Time: 01/10/2019 19:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES AVERY, DAVARIA, C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER, A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: MCFADDEN

DATE/TIME OF TEST: 12/25/2018 23:46:54

Diagnostics Check	OK	23:40
Air Blank	0.000	23:40
Control Test	0.081	23:41
Air Blank	0.000	23:41
Subject Sample # 1	0.134	23:42
Breath Volume	1.136	23:42
Air Blank	0.000	23:42
Air Blank	0.000	23:44
Subject Sample # 2	0.132	23:45
Breath Volume	1.593	23:45
Air Blank	0.000	23:45
Control Test	0.081	23:46
Air Blank	0.000	23:46
Diagnostics Check	OK	23:46

Data Download Date/Time: 01/10/2019 19:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PAYNTER,H
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:39

SUBJECT'S LAST NAME: MALDONADO

DATE/TIME OF TEST: 12/26/2018 05:06:03

Diagnostics Check	OK	04:59
Air Blank	0.000	04:59
Control Test	0.081	05:00
Air Blank	0.000	05:00
Subject Sample # 1	0.117	05:01
Breath Volume	3.457	05:01
Air Blank	0.000	05:02
Air Blank	0.000	05:03
Subject Sample # 2	0.111	05:04
Breath Volume	1.414	05:04
Air Blank	0.000	05:05
Control Test	0.080	05:05
Air Blank	0.000	05:05
Diagnostics Check	OK	05:06

Data Download Date/Time: 01/10/2019 19:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES AVERY, DAVARIA, C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: MARTINEZ HERNANDEZ

DATE/TIME OF TEST: 12/27/2018 03:55:49

Diagnostics Check	OK	03:49
Air Blank	0.000	03:49
Control Test	0.080	03:50
Air Blank	0.000	03:50
Subject Sample # 1	0.238	03:51
Breath Volume	1.269	03:51
Air Blank	0.000	03:51
Air Blank	0.000	03:53
Subject Sample # 2	0.250	03:54
Breath Volume	2.562	03:54
Air Blank	0.000	03:54
Control Test	0.079	03:55
Air Blank	0.000	03:55
Diagnostics Check	OK	03:55

Data Download Date/Time: 01/10/2019 19:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MOREJON,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: HERRMANN

DATE/TIME OF TEST: 12/29/2018 02:55:46

Diagnostics Check	OK	02:48
Air Blank	0.000	02:49
Control Test	0.080	02:49
Air Blank	0.000	02:49
Subject Sample # 1	0.124	02:51
Breath Volume	2.667	02:51
Air Blank	0.000	02:51
Air Blank	0.000	02:53
Subject Sample # 2	0.127	02:54
Breath Volume	2.421	02:54
Air Blank	0.000	02:54
Control Test	0.080	02:55
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

Data Download Date/Time: 01/10/2019 19:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES AVERY, DAVARIA, C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MOREJON, A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: TORNA MORALES

DATE/TIME OF TEST: 12/30/2018 02:02:57

Diagnostics Check	OK	02:00	
Air Blank	0.000	02:00	
Control Test	0.080	02:00	
Air Blank	0.000	02:01	
Subject Sample	# 1	02:01	Subject Test Refused
Breath Volume	0.000	02:01	
Air Blank	0.000	02:02	
Control Test	0.080	02:02	
Air Blank	0.000	02:02	
Diagnostics Check	OK	02:02	

Data Download Date/Time: 01/10/2019 19:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: VEGA CRUZ

DATE/TIME OF TEST: 12/30/2018 03:45:12

Diagnostics Check	OK	03:38
Air Blank	0.000	03:38
Control Test	0.080	03:38
Air Blank	0.000	03:39
Subject Sample # 1	0.103	03:40
Breath Volume	1.937	03:40
Air Blank	0.000	03:40
Air Blank	0.000	03:42
Subject Sample # 2	0.099	03:43
Breath Volume	1.941	03:43
Air Blank	0.000	03:44
Control Test	0.080	03:44
Air Blank	0.000	03:45
Diagnostics Check	OK	03:45

Data Download Date/Time: 01/10/2019 19:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PECYLAK,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: UZCATEGUI

DATE/TIME OF TEST: 12/30/2018 04:45:42

Diagnostics Check	OK	04:39
Air Blank	0.000	04:39
Control Test	0.080	04:40
Air Blank	0.000	04:40
Subject Sample # 1	0.098	04:41
Breath Volume	2.027	04:41
Air Blank	0.000	04:41
Air Blank	0.000	04:43
Subject Sample # 2	0.085	04:44
Breath Volume	1.289	04:44
Air Blank	0.000	04:44
Control Test	0.080	04:45
Air Blank	0.000	04:45
Diagnostics Check	OK	04:45

Data Download Date/Time: 01/10/2019 19:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIMPKINS,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: JUDACEWSKI

DATE/TIME OF TEST: 12/31/2018 02:25:38

Diagnostics Check	OK	02:19	
Air Blank	0.000	02:19	
Control Test	0.081	02:19	
Air Blank	0.000	02:20	
Subject Sample	# 1 0.212	02:21	
Breath Volume	2.414	02:21	
Air Blank	0.000	02:22	
Air Blank	0.000	02:24	
Subject Sample	# 2	02:24	RFI Detect
Breath Volume	0.000	02:24	
Air Blank	0.000	02:24	
Control Test	0.079	02:25	
Air Blank	0.000	02:25	
Diagnostics Check	OK	02:25	

Data Download Date/Time: 01/10/2019 19:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIMPKINS,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: JUDACEWSKI

DATE/TIME OF TEST: 12/31/2018 02:34:10

Diagnostics Check	OK	02:26
Air Blank	0.000	02:27
Control Test	0.081	02:27
Air Blank	0.000	02:28
Subject Sample # 1	0.204	02:28
Breath Volume	1.414	02:28
Air Blank	0.000	02:29
Air Blank	0.000	02:31
Subject Sample # 2	0.208	02:32
Breath Volume	2.535	02:32
Air Blank	0.000	02:33
Control Test	0.079	02:33
Air Blank	0.000	02:34
Diagnostics Check	OK	02:34

Data Download Date/Time: 01/10/2019 19:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES AVERY, DAVARIA, C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD, J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 12/31/2018 04:16:14

Diagnostics Check	OK	04:10
Air Blank	0.000	04:10
Control Test	0.080	04:10
Air Blank	0.000	04:11
Subject Sample # 1	0.208	04:11
Breath Volume	2.359	04:11
Air Blank	0.000	04:12
Air Blank	0.000	04:14
Subject Sample # 2	0.209	04:14
Breath Volume	2.222	04:14
Air Blank	0.000	04:15
Control Test	0.080	04:15
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

Data Download Date/Time: 01/10/2019 19:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALFY,RACHEL,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: PALFY,RACHEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: MORALES BAUTISTA

DATE/TIME OF TEST: 12/01/2018 03:54:25

Diagnostics Check	OK	03:53	
Air Blank		03:53	RFI Detect
Air Blank	0.000	03:54	

Data Download Date/Time: 01/02/2019 16:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALFY,RACHEL,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: PALFY,RACHEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: MORALES BAUTISTA

DATE/TIME OF TEST: 12/01/2018 04:03:40

Diagnostics Check	OK	03:56
Air Blank	0.000	03:56
Control Test	0.079	03:56
Air Blank	0.000	03:57
Subject Sample # 1	0.123	03:59
Breath Volume	1.617	03:59
Air Blank	0.000	03:59
Air Blank	0.000	04:01
Subject Sample # 2	0.125	04:02
Breath Volume	1.765	04:02
Air Blank	0.000	04:02
Control Test	0.079	04:03
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time: 01/02/2019 16:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALFY,RACHEL,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: PALFY,RACHEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: CRUZ CRUZ

DATE/TIME OF TEST: 12/02/2018 02:38:13

Diagnostics Check	OK	02:34	
Air Blank	0.000	02:34	
Control Test	0.079	02:35	
Air Blank	0.000	02:35	
Subject Sample	# 1	02:36	RFI Detect
Breath Volume	0.832	02:36	
Air Blank	0.000	02:37	
Control Test	0.079	02:37	
Air Blank		02:37	RFI Detect
Air Blank	0.000	02:38	

Data Download Date/Time: 01/02/2019 16:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALFY,RACHEL,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: PALFY,RACHEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: ED DOUZI

DATE/TIME OF TEST: 12/07/2018 02:51:53

Diagnostics Check	OK	02:41	
Air Blank	0.000	02:41	
Control Test	0.080	02:42	
Air Blank	0.000	02:42	
Subject Sample	# 1	02:45	No Sample Provided
Breath Volume	0.000	02:45	
Air Blank	0.000	02:46	
Air Blank	0.000	02:48	
Subject Sample	# 2	02:50	Subject Test Refused
Breath Volume	0.000	02:50	
Air Blank	0.000	02:50	
Control Test	0.080	02:51	
Air Blank	0.000	02:51	
Diagnostics Check	OK	02:51	

Data Download Date/Time: 01/02/2019 16:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: CARSON

DATE/TIME OF TEST: 12/15/2018 03:09:09

Diagnostics Check	OK	03:03	
Air Blank	0.000	03:03	
Control Test	0.080	03:04	
Air Blank	0.000	03:04	
Subject Sample # 1	0.094	03:05	
Breath Volume	1.597	03:05	
Air Blank	0.000	03:06	
Air Blank		03:07	RFI Detect
Air Blank	0.000	03:08	
Control Test	0.080	03:08	
Air Blank	0.000	03:09	
Diagnostics Check	OK	03:09	

Data Download Date/Time: 01/02/2019 16:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: CARSON

DATE/TIME OF TEST: 12/15/2018 03:17:15

Diagnostics Check	OK	03:10
Air Blank	0.000	03:11
Control Test	0.079	03:11
Air Blank	0.000	03:12
Subject Sample # 1	0.093	03:12
Breath Volume	1.910	03:12
Air Blank	0.000	03:13
Air Blank	0.000	03:15
Subject Sample # 2	0.090	03:15
Breath Volume	1.953	03:15
Air Blank	0.000	03:16
Control Test	0.079	03:16
Air Blank	0.000	03:17
Diagnostics Check	OK	03:17

Data Download Date/Time: 01/02/2019 16:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: US US

DATE/TIME OF TEST: 12/16/2018 01:24:25

Diagnostics Check	OK	01:18
Air Blank	0.000	01:18
Control Test	0.079	01:18
Air Blank	0.000	01:19
Subject Sample # 1	0.232	01:19
Breath Volume	1.441	01:19
Air Blank	0.000	01:20
Air Blank	0.000	01:22
Subject Sample # 2	0.230	01:22
Breath Volume	1.339	01:22
Air Blank	0.000	01:23
Control Test	0.080	01:23
Air Blank	0.000	01:24
Diagnostics Check	OK	01:24

Data Download Date/Time: 01/02/2019 16:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALFY,RACHEL,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: PALFY,RACHEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: CASTRO MENDOZA

DATE/TIME OF TEST: 12/22/2018 01:57:57

Diagnostics Check	OK	01:50
Air Blank	0.000	01:51
Control Test	0.080	01:51
Air Blank	0.000	01:51
Subject Sample # 1	0.211	01:53
Breath Volume	1.667	01:53
Air Blank	0.000	01:53
Air Blank	0.000	01:55
Subject Sample # 2	0.217	01:56
Breath Volume	1.789	01:56
Air Blank	0.000	01:57
Control Test	0.080	01:57
Air Blank	0.000	01:57
Diagnostics Check	OK	01:57

Data Download Date/Time: 01/02/2019 16:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALFY,RACHEL,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: PALFY,RACHEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: CHAVARRIA

DATE/TIME OF TEST: 12/26/2018 03:08:20

Diagnostics Check	OK	03:01
Air Blank	0.000	03:01
Control Test	0.080	03:01
Air Blank	0.000	03:02
Subject Sample # 1	0.191	03:03
Breath Volume	2.250	03:03
Air Blank	0.000	03:04
Air Blank	0.000	03:06
Subject Sample # 2	0.178	03:06
Breath Volume	1.878	03:06
Air Blank	0.000	03:07
Control Test	0.078	03:07
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

Data Download Date/Time: 01/02/2019 16:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: VASQUEZ

DATE/TIME OF TEST: 12/29/2018 03:43:07

Diagnostics Check	OK	03:36
Air Blank	0.000	03:37
Control Test	0.079	03:37
Air Blank	0.000	03:38
Subject Sample # 1	0.228	03:38
Breath Volume	2.046	03:38
Air Blank	0.000	03:39
Air Blank	0.000	03:41
Subject Sample # 2	0.235	03:41
Breath Volume	1.769	03:41
Air Blank	0.000	03:42
Control Test	0.079	03:42
Air Blank	0.000	03:43
Diagnostics Check	OK	03:43

Data Download Date/Time: 01/02/2019 16:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,JUSTIN,M
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: DOOLIN,EMILY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 11/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: PROKOP

DATE/TIME OF TEST: 12/17/2018 04:48:08

Diagnostics Check	OK	04:41
Air Blank	0.000	04:42
Control Test	0.080	04:42
Air Blank	0.000	04:42
Subject Sample # 1	0.124	04:43
Breath Volume	2.757	04:43
Air Blank	0.000	04:44
Air Blank	0.000	04:46
Subject Sample # 2	0.121	04:46
Breath Volume	1.949	04:46
Air Blank	0.000	04:47
Control Test	0.080	04:47
Air Blank	0.000	04:48
Diagnostics Check	OK	04:48

Data Download Date/Time: 12/19/2018 13:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FITCHETT, ADAM, R
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SWAIN, ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: VALDEZ

DATE/TIME OF TEST: 12/01/2018 04:41:56

Diagnostics Check	OK	04:35
Air Blank	0.000	04:36
Control Test	0.080	04:36
Air Blank	0.000	04:37
Subject Sample # 1	0.205	04:37
Breath Volume	1.511	04:37
Air Blank	0.000	04:38
Air Blank	0.000	04:39
Subject Sample # 2	0.195	04:40
Breath Volume	2.960	04:40
Air Blank	0.000	04:41
Control Test	0.080	04:41
Air Blank	0.000	04:41
Diagnostics Check	OK	04:41

Data Download Date/Time: 12/06/2018 09:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RACER,JACOB,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: RACER,JACOB
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: MCCURDY

DATE/TIME OF TEST: 12/06/2018 04:16:56

Diagnostics Check	OK	04:05	
Air Blank	0.000	04:06	
Control Test	0.082	04:06	
Air Blank	0.000	04:06	
Subject Sample	# 1	04:09	No Sample Provided
Breath Volume	0.000	04:09	
Air Blank	0.000	04:10	
Air Blank	0.000	04:12	
Subject Sample	# 2	04:15	No Sample Provided
Breath Volume	0.000	04:15	
Air Blank	0.000	04:16	
Control Test	0.082	04:16	
Air Blank	0.000	04:16	
Diagnostics Check	OK	04:16	

Data Download Date/Time: 12/06/2018 09:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER, TYLER, J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FOWLER, TYLER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 11/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: SCHAFFER

DATE/TIME OF TEST: 12/01/2018 19:40:13

Diagnostics Check	OK	19:31	
Air Blank	0.000	19:32	
Control Test	0.077	19:32	
Air Blank	0.000	19:33	
Subject Sample	# 1	19:36	No Sample Provided
Breath Volume	0.000	19:36	
Air Blank	0.000	19:36	
Air Blank	0.000	19:38	
Subject Sample	# 2	19:38	Subject Test Refused
Breath Volume	0.000	19:38	
Air Blank	0.000	19:39	
Control Test	0.077	19:39	
Air Blank	0.000	19:40	
Diagnostics Check	OK	19:40	

Data Download Date/Time: 12/28/2018 09:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER, TYLER, J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FOWLER, TYLER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 11/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 12/14/2018 00:28:25

Diagnostics Check	OK	00:22
Air Blank	0.000	00:22
Control Test	0.078	00:23
Air Blank	0.000	00:23
Subject Sample # 1	0.205	00:24
Breath Volume	3.023	00:24
Air Blank	0.000	00:24
Air Blank	0.000	00:26
Subject Sample # 2	0.202	00:26
Breath Volume	2.367	00:26
Air Blank	0.000	00:27
Control Test	0.077	00:27
Air Blank	0.000	00:28
Diagnostics Check	OK	00:28

Data Download Date/Time: 12/28/2018 09:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER, TYLER,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FOWLER, TYLER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 11/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 12/16/2018 00:16:38

Diagnostics Check	OK	00:10
Air Blank	0.000	00:10
Control Test	0.077	00:10
Air Blank	0.000	00:11
Subject Sample # 1	0.130	00:12
Breath Volume	3.746	00:12
Air Blank	0.000	00:12
Air Blank	0.000	00:14
Subject Sample # 2	0.133	00:15
Breath Volume	3.628	00:15
Air Blank	0.000	00:15
Control Test	0.077	00:16
Air Blank	0.000	00:16
Diagnostics Check	OK	00:16

Data Download Date/Time: 12/28/2018 09:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANDERSON,C
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: FRITZ

DATE/TIME OF TEST: 12/01/2018 00:21:33

Diagnostics Check	OK	00:14
Air Blank	0.000	00:15
Control Test	0.082	00:15
Air Blank	0.000	00:15
Subject Sample # 1	0.118	00:16
Breath Volume	2.281	00:16
Air Blank	0.000	00:17
Air Blank	0.000	00:19
Subject Sample # 2	0.115	00:20
Breath Volume	2.281	00:20
Air Blank	0.000	00:20
Control Test	0.081	00:20
Air Blank	0.000	00:21
Diagnostics Check	OK	00:21

Data Download Date/Time: 12/18/2018 17:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER, A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: DAVIDSON

DATE/TIME OF TEST: 12/01/2018 04:50:29

Diagnostics Check	OK	04:43
Air Blank	0.000	04:44
Control Test	0.081	04:44
Air Blank	0.000	04:45
Subject Sample # 1	0.160	04:45
Breath Volume	1.562	04:45
Air Blank	0.000	04:46
Air Blank	0.000	04:48
Subject Sample # 2	0.153	04:48
Breath Volume	1.113	04:48
Air Blank	0.000	04:49
Control Test	0.081	04:49
Air Blank	0.000	04:50
Diagnostics Check	OK	04:50

Data Download Date/Time: 12/18/2018 17:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PECYLAK,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:14

SUBJECT'S LAST NAME: KRAUSE

DATE/TIME OF TEST: 12/01/2018 05:42:38

Diagnostics Check	OK	05:34
Air Blank	0.000	05:34
Control Test	0.081	05:35
Air Blank	0.000	05:35
Subject Sample # 1	0.154	05:37
Breath Volume	1.390	05:37
Air Blank	0.000	05:37
Air Blank	0.000	05:39
Subject Sample # 2	0.151	05:41
Breath Volume	1.304	05:41
Air Blank	0.000	05:41
Control Test	0.081	05:42
Air Blank	0.000	05:42
Diagnostics Check	OK	05:42

Data Download Date/Time: 12/18/2018 17:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,REBECCA,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:11

SUBJECT'S LAST NAME: PARIS

DATE/TIME OF TEST: 12/01/2018 21:45:48

Diagnostics Check	OK	21:39
Air Blank	0.000	21:39
Control Test	0.081	21:40
Air Blank	0.000	21:40
Subject Sample # 1	0.193	21:41
Breath Volume	1.765	21:41
Air Blank	0.000	21:42
Air Blank	0.000	21:43
Subject Sample # 2	0.193	21:44
Breath Volume	1.226	21:44
Air Blank	0.000	21:44
Control Test	0.081	21:45
Air Blank	0.000	21:45
Diagnostics Check	OK	21:45

Data Download Date/Time: 12/18/2018 17:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: MERCADO GUARDADO

DATE/TIME OF TEST: 12/01/2018 23:49:26

Diagnostics Check	OK	23:42
Air Blank	0.000	23:42
Control Test	0.081	23:43
Air Blank	0.000	23:43
Subject Sample # 1	0.143	23:44
Breath Volume	1.839	23:44
Air Blank	0.000	23:45
Air Blank	0.000	23:47
Subject Sample # 2	0.131	23:47
Breath Volume	1.234	23:47
Air Blank	0.000	23:48
Control Test	0.081	23:48
Air Blank	0.000	23:49
Diagnostics Check	OK	23:49

Data Download Date/Time: 12/18/2018 17:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,REBECCA,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ARCHAMBEAU,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: MARINE

DATE/TIME OF TEST: 12/02/2018 02:33:05

Diagnostics Check	OK	02:26
Air Blank	0.000	02:27
Control Test	0.081	02:27
Air Blank	0.000	02:27
Subject Sample # 1	0.161	02:28
Breath Volume	1.785	02:28
Air Blank	0.000	02:29
Air Blank	0.000	02:31
Subject Sample # 2	0.166	02:31
Breath Volume	1.964	02:31
Air Blank	0.000	02:32
Control Test	0.081	02:32
Air Blank	0.000	02:33
Diagnostics Check	OK	02:33

Data Download Date/Time: 12/18/2018 17:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: WITZIGMAN

DATE/TIME OF TEST: 12/02/2018 02:54:44

Diagnostics Check	OK	02:48
Air Blank	0.000	02:48
Control Test	0.081	02:49
Air Blank	0.000	02:49
Subject Sample # 1	0.136	02:50
Breath Volume	2.136	02:50
Air Blank	0.000	02:50
Air Blank	0.000	02:52
Subject Sample # 2	0.132	02:53
Breath Volume	1.476	02:53
Air Blank	0.000	02:53
Control Test	0.082	02:54
Air Blank	0.000	02:54
Diagnostics Check	OK	02:54

Data Download Date/Time: 12/18/2018 17:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: RAGANS

DATE/TIME OF TEST: 12/02/2018 04:26:18

Diagnostics Check	OK	04:20
Air Blank	0.000	04:20
Control Test	0.080	04:20
Air Blank	0.000	04:21
Subject Sample # 1	0.032	04:21
Breath Volume	2.808	04:21
Air Blank	0.000	04:22
Air Blank	0.000	04:24
Subject Sample # 2	0.028	04:24
Breath Volume	1.808	04:24
Air Blank	0.000	04:25
Control Test	0.081	04:25
Air Blank	0.000	04:26
Diagnostics Check	OK	04:26

Data Download Date/Time: 12/18/2018 17:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,REBECCA,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DRISCOLL,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: RAMOS MELENDEZ

DATE/TIME OF TEST: 12/02/2018 22:46:36

Diagnostics Check	OK	22:40
Air Blank	0.000	22:40
Control Test	0.081	22:41
Air Blank	0.000	22:41
Subject Sample # 1	0.000	22:42
Breath Volume	1.863	22:42
Air Blank	0.000	22:42
Air Blank	0.000	22:44
Subject Sample # 2	0.000	22:45
Breath Volume	2.808	22:45
Air Blank	0.000	22:45
Control Test	0.082	22:46
Air Blank	0.000	22:46
Diagnostics Check	OK	22:46

Data Download Date/Time: 12/18/2018 17:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,REBECCA,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: SUSSMAN

DATE/TIME OF TEST: 12/02/2018 23:32:01

Diagnostics Check	OK	23:25
Air Blank	0.000	23:25
Control Test	0.081	23:26
Air Blank	0.000	23:26
Subject Sample # 1	0.243	23:27
Breath Volume	2.625	23:27
Air Blank	0.000	23:28
Air Blank	0.000	23:29
Subject Sample # 2	0.243	23:30
Breath Volume	3.363	23:30
Air Blank	0.000	23:31
Control Test	0.081	23:31
Air Blank	0.000	23:31
Diagnostics Check	OK	23:32

Data Download Date/Time: 12/18/2018 17:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FISHER,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:57

SUBJECT'S LAST NAME: HEGLER

DATE/TIME OF TEST: 12/03/2018 19:26:08

Diagnostics Check	OK	19:18
Air Blank	0.000	19:19
Control Test	0.082	19:19
Air Blank	0.000	19:20
Subject Sample # 1	0.031	19:21
Breath Volume	1.386	19:21
Air Blank	0.000	19:22
Air Blank	0.000	19:24
Subject Sample # 2	0.030	19:24
Breath Volume	1.433	19:24
Air Blank	0.000	19:25
Control Test	0.082	19:25
Air Blank	0.000	19:26
Diagnostics Check	OK	19:26

Data Download Date/Time: 12/18/2018 17:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SHEINBERG,T
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: BOGHICH

DATE/TIME OF TEST: 12/05/2018 03:33:11

Diagnostics Check	OK	03:26
Air Blank	0.000	03:27
Control Test	0.082	03:27
Air Blank	0.000	03:27
Subject Sample # 1	0.210	03:28
Breath Volume	2.542	03:28
Air Blank	0.000	03:29
Air Blank	0.000	03:31
Subject Sample # 2	0.207	03:31
Breath Volume	1.882	03:31
Air Blank	0.000	03:32
Control Test	0.082	03:32
Air Blank	0.000	03:33
Diagnostics Check	OK	03:33

Data Download Date/Time: 12/18/2018 17:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MOREJON,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: BOYD

DATE/TIME OF TEST: 12/05/2018 22:25:32

Diagnostics Check	OK	22:16
Air Blank	0.000	22:17
Control Test	0.082	22:17
Air Blank	0.000	22:18
Subject Sample # 1	0.092	22:20
Breath Volume	1.792	22:20
Air Blank	0.000	22:20
Air Blank	0.000	22:22
Subject Sample # 2	0.091	22:24
Breath Volume	1.945	22:24
Air Blank	0.000	22:24
Control Test	0.081	22:24
Air Blank	0.000	22:25
Diagnostics Check	OK	22:25

Data Download Date/Time: 12/18/2018 17:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DRISCOLL, B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: CROCKER

DATE/TIME OF TEST: 12/06/2018 05:27:15

Diagnostics Check	OK	05:15	
Air Blank	0.000	05:16	
Control Test	0.082	05:16	
Air Blank	0.000	05:16	
Subject Sample	# 1 0.115	05:20	Volume Not Met
Breath Volume	0.042	05:20	
Air Blank	0.000	05:20	
Air Blank	0.000	05:22	
Subject Sample	# 2 0.123	05:25	Volume Not Met
Breath Volume	0.023	05:25	
Air Blank	0.000	05:26	
Control Test	0.083	05:26	
Air Blank	0.000	05:27	
Diagnostics Check	OK	05:27	

Data Download Date/Time: 12/18/2018 17:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DRISCOLL, B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: CROCKER

DATE/TIME OF TEST: 12/06/2018 05:39:47

Diagnostics Check	OK	05:28	
Air Blank	0.000	05:28	
Control Test	0.081	05:28	
Air Blank	0.000	05:29	
Subject Sample	# 1 0.126	05:32	Volume Not Met
Breath Volume	0.070	05:32	
Air Blank	0.000	05:33	
Air Blank	0.000	05:35	
Subject Sample	# 2 0.112	05:38	Volume Not Met
Breath Volume	0.019	05:38	
Air Blank	0.000	05:38	
Control Test	0.082	05:39	
Air Blank	0.000	05:39	
Diagnostics Check	OK	05:39	

Data Download Date/Time: 12/18/2018 17:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: ARROYO

DATE/TIME OF TEST: 12/07/2018 04:13:10

Diagnostics Check	OK	04:06
Air Blank	0.000	04:06
Control Test	0.082	04:06
Air Blank	0.000	04:07
Subject Sample # 1	0.213	04:07
Breath Volume	1.250	04:07
Air Blank	0.000	04:08
Air Blank	0.000	04:10
Subject Sample # 2	0.210	04:11
Breath Volume	2.136	04:11
Air Blank	0.000	04:12
Control Test	0.082	04:12
Air Blank	0.000	04:13
Diagnostics Check	OK	04:13

Data Download Date/Time: 12/18/2018 17:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: MASON

DATE/TIME OF TEST: 12/08/2018 00:43:13

Diagnostics Check	OK	00:36
Air Blank	0.000	00:36
Control Test	0.082	00:36
Air Blank	0.000	00:37
Subject Sample # 1	0.183	00:38
Breath Volume	1.957	00:38
Air Blank	0.000	00:39
Air Blank	0.000	00:41
Subject Sample # 2	0.176	00:41
Breath Volume	1.386	00:41
Air Blank	0.000	00:42
Control Test	0.081	00:42
Air Blank	0.000	00:43
Diagnostics Check	OK	00:43

Data Download Date/Time: 12/18/2018 17:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: CIGRAND

DATE/TIME OF TEST: 12/08/2018 01:14:51

Diagnostics Check	OK	01:00	
Air Blank	0.000	01:01	
Control Test	0.082	01:01	
Air Blank	0.000	01:01	
Subject Sample	# 1 0.157	01:04	
Breath Volume	1.679	01:04	
Air Blank	0.000	01:05	
Air Blank	0.000	01:06	
Subject Sample	# 2 0.138	01:10	Volume Not Met
Breath Volume	0.792	01:10	
Air Blank	0.000	01:10	
Air Blank	0.000	01:12	
Subject Sample	# 3 0.164	01:13	
Breath Volume	2.042	01:13	
Air Blank	0.000	01:13	
Control Test	0.081	01:14	
Air Blank	0.000	01:14	
Diagnostics Check	OK	01:14	

Data Download Date/Time: 12/18/2018 17:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: RIVERA SANTIAGO

DATE/TIME OF TEST: 12/08/2018 04:26:59

Diagnostics Check	OK	04:20
Air Blank	0.000	04:20
Control Test	0.081	04:21
Air Blank	0.000	04:21
Subject Sample # 1	0.122	04:22
Breath Volume	1.187	04:22
Air Blank	0.000	04:22
Air Blank	0.000	04:24
Subject Sample # 2	0.127	04:25
Breath Volume	1.300	04:25
Air Blank	0.000	04:26
Control Test	0.081	04:26
Air Blank	0.000	04:26
Diagnostics Check	OK	04:26

Data Download Date/Time: 12/18/2018 17:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: WASHINGTON

DATE/TIME OF TEST: 12/08/2018 04:42:36

Diagnostics Check	OK	04:36
Air Blank	0.000	04:36
Control Test	0.081	04:37
Air Blank	0.000	04:37
Subject Sample # 1	0.133	04:38
Breath Volume	2.976	04:38
Air Blank	0.000	04:38
Air Blank	0.000	04:40
Subject Sample # 2	0.139	04:41
Breath Volume	4.410	04:41
Air Blank	0.000	04:41
Control Test	0.082	04:42
Air Blank	0.000	04:42
Diagnostics Check	OK	04:42

Data Download Date/Time: 12/18/2018 17:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: MULVIHILL

DATE/TIME OF TEST: 12/09/2018 01:30:20

Diagnostics Check	OK	01:23
Air Blank	0.000	01:23
Control Test	0.082	01:23
Air Blank	0.000	01:24
Subject Sample # 1	0.253	01:25
Breath Volume	1.875	01:25
Air Blank	0.000	01:26
Air Blank	0.000	01:27
Subject Sample # 2	0.253	01:28
Breath Volume	1.386	01:28
Air Blank	0.000	01:29
Control Test	0.082	01:29
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

Data Download Date/Time: 12/18/2018 17:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FLORES, C
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: STEEVES

DATE/TIME OF TEST: 12/09/2018 02:00:17

Diagnostics Check	OK	01:54
Air Blank	0.000	01:54
Control Test	0.082	01:54
Air Blank	0.000	01:55
Subject Sample # 1	0.052	01:55
Breath Volume	3.050	01:55
Air Blank	0.000	01:56
Air Blank	0.000	01:58
Subject Sample # 2	0.053	01:58
Breath Volume	3.156	01:58
Air Blank	0.000	01:59
Control Test	0.082	01:59
Air Blank	0.000	02:00
Diagnostics Check	OK	02:00

Data Download Date/Time: 12/18/2018 17:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRIGHTWELL, R
ARRESTING OFFICER AGENCY: U of So FL Campus PD
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 12/09/2018 02:45:35

Diagnostics Check	OK	02:39
Air Blank	0.000	02:40
Control Test	0.082	02:40
Air Blank	0.000	02:40
Subject Sample # 1	0.023	02:41
Breath Volume	2.257	02:41
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.023	02:44
Breath Volume	1.328	02:44
Air Blank	0.000	02:44
Control Test	0.082	02:44
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 12/18/2018 17:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FISHER,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 12/09/2018 04:17:52

Diagnostics Check	OK	04:10
Air Blank	0.000	04:10
Control Test	0.081	04:11
Air Blank	0.000	04:11
Subject Sample # 1	0.138	04:13
Breath Volume	1.488	04:13
Air Blank	0.000	04:14
Air Blank	0.000	04:15
Subject Sample # 2	0.139	04:16
Breath Volume	1.187	04:16
Air Blank	0.000	04:16
Control Test	0.081	04:17
Air Blank	0.000	04:17
Diagnostics Check	OK	04:17

Data Download Date/Time: 12/18/2018 17:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: GARZA JARAMILLO

DATE/TIME OF TEST: 12/10/2018 00:06:27

Diagnostics Check	OK	23:59
Air Blank	0.000	00:00
Control Test	0.081	00:00
Air Blank	0.000	00:01
Subject Sample # 1	0.202	00:02
Breath Volume	2.375	00:02
Air Blank	0.000	00:02
Air Blank	0.000	00:04
Subject Sample # 2	0.183	00:04
Breath Volume	1.632	00:04
Air Blank	0.000	00:05
Control Test	0.081	00:05
Air Blank	0.000	00:06
Diagnostics Check	OK	00:06

Data Download Date/Time: 12/18/2018 17:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SCHILLER, E
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 12/10/2018 01:34:52

Diagnostics Check	OK	01:28
Air Blank	0.000	01:29
Control Test	0.081	01:29
Air Blank	0.000	01:29
Subject Sample # 1	0.112	01:30
Breath Volume	2.011	01:30
Air Blank	0.000	01:31
Air Blank	0.000	01:32
Subject Sample # 2	0.108	01:33
Breath Volume	1.578	01:33
Air Blank	0.000	01:33
Control Test	0.081	01:34
Air Blank	0.000	01:34
Diagnostics Check	OK	01:34

Data Download Date/Time: 12/18/2018 17:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: EDGE, S
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: CRAMER

DATE/TIME OF TEST: 12/10/2018 22:49:20

Diagnostics Check	OK	22:42
Air Blank	0.000	22:42
Control Test	0.082	22:43
Air Blank	0.000	22:43
Subject Sample # 1	0.108	22:44
Breath Volume	2.632	22:44
Air Blank	0.000	22:45
Air Blank	0.000	22:46
Subject Sample # 2	0.106	22:47
Breath Volume	3.132	22:47
Air Blank	0.000	22:48
Control Test	0.081	22:48
Air Blank	0.000	22:49
Diagnostics Check	OK	22:49

Data Download Date/Time: 12/18/2018 17:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MEDINA,P
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: ROUNTREE

DATE/TIME OF TEST: 12/11/2018 00:41:04

Diagnostics Check	OK	00:37	
Air Blank	0.000	00:37	
Control Test	0.082	00:38	
Air Blank	0.000	00:38	
Subject Sample	# 1	00:39	Subject Test Refused
Breath Volume	0.000	00:39	
Air Blank	0.000	00:40	
Control Test	0.083	00:40	
Air Blank	0.000	00:40	
Diagnostics Check	OK	00:41	

Data Download Date/Time: 12/18/2018 17:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PAYNTER,H
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: PERSHAD

DATE/TIME OF TEST: 12/11/2018 05:03:48

Diagnostics Check	OK	04:57
Air Blank	0.000	04:57
Control Test	0.082	04:57
Air Blank	0.000	04:58
Subject Sample # 1	0.235	04:59
Breath Volume	1.132	04:59
Air Blank	0.000	04:59
Air Blank	0.000	05:01
Subject Sample # 2	0.251	05:02
Breath Volume	1.515	05:02
Air Blank	0.000	05:02
Control Test	0.081	05:03
Air Blank	0.000	05:03
Diagnostics Check	OK	05:03

Data Download Date/Time: 12/18/2018 17:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: COSTA,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:39

SUBJECT'S LAST NAME: WYNNE

DATE/TIME OF TEST: 12/11/2018 21:06:10

Diagnostics Check	OK	20:59
Air Blank	0.000	20:59
Control Test	0.082	20:59
Air Blank	0.000	21:00
Subject Sample # 1	0.086	21:01
Breath Volume	1.910	21:01
Air Blank	0.000	21:01
Air Blank	0.000	21:03
Subject Sample # 2	0.083	21:04
Breath Volume	1.457	21:04
Air Blank	0.000	21:05
Control Test	0.082	21:05
Air Blank	0.000	21:06
Diagnostics Check	OK	21:06

Data Download Date/Time: 12/18/2018 17:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PECYLAK,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: ROSA-PLAZA

DATE/TIME OF TEST: 12/12/2018 04:53:43

Diagnostics Check	OK	04:45	
Air Blank	0.000	04:45	
Control Test	0.083	04:45	
Air Blank	0.000	04:46	
Subject Sample	# 1	04:49	No Sample Provided
Breath Volume	0.000	04:49	
Air Blank	0.000	04:50	
Air Blank	0.000	04:52	
Subject Sample	# 2	04:52	Subject Test Refused
Breath Volume	0.000	04:52	
Air Blank	0.000	04:52	
Control Test	0.082	04:53	
Air Blank	0.000	04:53	
Diagnostics Check	OK	04:53	

Data Download Date/Time: 12/18/2018 17:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: JOINER

DATE/TIME OF TEST: 12/12/2018 23:34:17

Diagnostics Check	OK	23:27
Air Blank	0.000	23:27
Control Test	0.082	23:27
Air Blank	0.000	23:28
Subject Sample # 1	0.247	23:29
Breath Volume	2.093	23:29
Air Blank	0.000	23:29
Air Blank	0.000	23:31
Subject Sample # 2	0.231	23:32
Breath Volume	1.296	23:32
Air Blank	0.000	23:33
Control Test	0.081	23:33
Air Blank	0.000	23:34
Diagnostics Check	OK	23:34

Data Download Date/Time: 12/18/2018 17:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,REBECCA,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: STEPHENS

DATE/TIME OF TEST: 12/13/2018 02:17:39

Diagnostics Check	OK	02:10
Air Blank	0.000	02:11
Control Test	0.082	02:11
Air Blank	0.000	02:11
Subject Sample # 1	0.106	02:13
Breath Volume	4.527	02:13
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample # 2	0.108	02:16
Breath Volume	4.136	02:16
Air Blank	0.000	02:16
Control Test	0.082	02:17
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

Data Download Date/Time: 12/18/2018 17:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: THORNE,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:47

SUBJECT'S LAST NAME: ANTONIO JOSE

DATE/TIME OF TEST: 12/13/2018 21:21:47

Diagnostics Check	OK	21:10	
Air Blank	0.000	21:10	
Control Test	0.082	21:10	
Air Blank	0.000	21:11	
Subject Sample	# 1 0.152	21:14	Volume Not Met
Breath Volume	0.054	21:14	
Air Blank	0.000	21:15	
Air Blank	0.000	21:16	
Subject Sample	# 2 0.174	21:20	Volume Not Met
Breath Volume	0.082	21:20	
Air Blank	0.000	21:20	
Control Test	0.084	21:21	
Air Blank	0.000	21:21	
Diagnostics Check	OK	21:21	

Data Download Date/Time: 12/18/2018 17:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 12/14/2018 02:38:09

Diagnostics Check	OK	02:30
Air Blank	0.000	02:31
Control Test	0.081	02:31
Air Blank	0.000	02:31
Subject Sample # 1	0.135	02:33
Breath Volume	1.988	02:33
Air Blank	0.000	02:33
Air Blank	0.000	02:35
Subject Sample # 2	0.148	02:36
Breath Volume	2.265	02:36
Air Blank	0.000	02:37
Control Test	0.082	02:37
Air Blank	0.000	02:38
Diagnostics Check	OK	02:38

Data Download Date/Time: 12/18/2018 17:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: JUDSON

DATE/TIME OF TEST: 12/15/2018 03:33:04

Diagnostics Check	OK	03:25
Air Blank	0.000	03:26
Control Test	0.082	03:26
Air Blank	0.000	03:27
Subject Sample # 1	0.135	03:28
Breath Volume	2.023	03:28
Air Blank	0.000	03:29
Air Blank	0.000	03:30
Subject Sample # 2	0.129	03:31
Breath Volume	2.089	03:31
Air Blank	0.000	03:32
Control Test	0.082	03:32
Air Blank	0.000	03:32
Diagnostics Check	OK	03:33

Data Download Date/Time: 12/18/2018 17:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ARCHAMBEAU, M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: LACROSS

DATE/TIME OF TEST: 12/15/2018 04:24:58

Diagnostics Check	OK	04:18
Air Blank	0.000	04:19
Control Test	0.082	04:19
Air Blank	0.000	04:20
Subject Sample # 1	0.197	04:20
Breath Volume	3.167	04:20
Air Blank	0.000	04:21
Air Blank	0.000	04:23
Subject Sample # 2	0.192	04:23
Breath Volume	1.500	04:23
Air Blank	0.000	04:24
Control Test	0.081	04:24
Air Blank	0.000	04:24
Diagnostics Check	OK	04:24

Data Download Date/Time: 12/18/2018 17:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DRAKE,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 13:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:29

SUBJECT'S LAST NAME: PURCE

DATE/TIME OF TEST: 12/15/2018 14:55:56

Diagnostics Check	OK	14:49
Air Blank	0.000	14:50
Control Test	0.082	14:50
Air Blank	0.000	14:50
Subject Sample # 1	0.000	14:51
Breath Volume	1.710	14:51
Air Blank	0.000	14:51
Air Blank	0.000	14:53
Subject Sample # 2	0.000	14:54
Breath Volume	2.167	14:54
Air Blank	0.000	14:55
Control Test	0.082	14:55
Air Blank	0.000	14:55
Diagnostics Check	OK	14:55

Data Download Date/Time: 12/18/2018 17:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: COSTA,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 12/16/2018 00:42:30

Diagnostics Check	OK	00:35
Air Blank	0.000	00:35
Control Test	0.082	00:36
Air Blank	0.000	00:36
Subject Sample # 1	0.230	00:37
Breath Volume	1.902	00:37
Air Blank	0.000	00:38
Air Blank	0.000	00:40
Subject Sample # 2	0.230	00:40
Breath Volume	1.949	00:40
Air Blank	0.000	00:41
Control Test	0.082	00:41
Air Blank	0.000	00:42
Diagnostics Check	OK	00:42

Data Download Date/Time: 12/18/2018 17:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, REBECCA, J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG, G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: ALI

DATE/TIME OF TEST: 12/16/2018 04:55:30

Diagnostics Check	OK	04:43	
Air Blank	0.000	04:43	
Control Test	0.082	04:43	
Air Blank	0.000	04:44	
Subject Sample # 1	0.147	04:45	
Breath Volume	2.515	04:45	
Air Blank	0.000	04:46	
Air Blank	0.000	04:48	
Subject Sample # 2		04:51	No Sample Provided
Breath Volume	0.000	04:51	
Air Blank	0.000	04:51	
Air Blank	0.000	04:53	
Subject Sample # 3		04:54	Subject Test Refused
Breath Volume	0.000	04:54	
Air Blank	0.000	04:54	
Control Test	0.081	04:54	
Air Blank	0.000	04:55	
Diagnostics Check	OK	04:55	

Data Download Date/Time: 12/18/2018 17:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: ROBIOU

DATE/TIME OF TEST: 12/18/2018 02:24:35

Diagnostics Check	OK	02:16
Air Blank	0.000	02:17
Control Test	0.082	02:17
Air Blank	0.000	02:18
Subject Sample # 1	0.189	02:19
Breath Volume	1.296	02:19
Air Blank	0.000	02:19
Air Blank	0.000	02:21
Subject Sample # 2	0.170	02:23
Breath Volume	1.339	02:23
Air Blank	0.000	02:23
Control Test	0.082	02:23
Air Blank	0.000	02:24
Diagnostics Check	OK	02:24

Data Download Date/Time: 12/18/2018 17:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FLORES,C
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 05:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:51

SUBJECT'S LAST NAME: CASTRO-CABRERA

DATE/TIME OF TEST: 12/18/2018 06:19:00

Diagnostics Check	OK	06:11
Air Blank	0.000	06:12
Control Test	0.082	06:12
Air Blank	0.000	06:13
Subject Sample # 1	0.060	06:14
Breath Volume	1.332	06:14
Air Blank	0.000	06:14
Air Blank	0.000	06:16
Subject Sample # 2	0.062	06:17
Breath Volume	1.710	06:17
Air Blank	0.000	06:18
Control Test	0.082	06:18
Air Blank	0.000	06:18
Diagnostics Check	OK	06:19

Data Download Date/Time: 12/18/2018 17:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEADOWS,GEOFFREY,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CROFT,BRETT
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 11/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A3
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: VENDLANDS

DATE/TIME OF TEST: 12/03/2018 23:38:16

Diagnostics Check	OK	23:30
Air Blank	0.000	23:31
Control Test	0.080	23:31
Air Blank	0.000	23:31
Subject Sample # 1	0.170	23:33
Breath Volume	2.433	23:33
Air Blank	0.000	23:34
Air Blank	0.000	23:36
Subject Sample # 2	0.167	23:36
Breath Volume	2.082	23:36
Air Blank	0.000	23:37
Control Test	0.080	23:37
Air Blank	0.000	23:38
Diagnostics Check	OK	23:38

Data Download Date/Time: 12/10/2018 20:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEADOWS,GEOFFREY,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CROFT,BRETT
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 11/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A3
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: GRANT

DATE/TIME OF TEST: 12/04/2018 21:00:14

Diagnostics Check	OK	20:55	
Air Blank	0.000	20:55	
Control Test	0.081	20:56	
Air Blank	0.000	20:56	
Subject Sample	# 1	20:58	Subject Test Refused
Breath Volume	0.000	20:58	
Air Blank	0.000	20:59	
Control Test	0.081	20:59	
Air Blank	0.000	21:00	
Diagnostics Check	OK	21:00	

Data Download Date/Time: 12/10/2018 20:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLLITO,BRAD,R
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: WILSON,KYLE
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 11/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A3
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: OSUNA GARCIA

DATE/TIME OF TEST: 12/07/2018 04:39:54

Diagnostics Check	OK	04:33
Air Blank	0.000	04:33
Control Test	0.080	04:34
Air Blank	0.000	04:34
Subject Sample # 1	0.245	04:35
Breath Volume	3.593	04:35
Air Blank	0.000	04:36
Air Blank	0.000	04:37
Subject Sample # 2	0.244	04:38
Breath Volume	3.742	04:38
Air Blank	0.000	04:38
Control Test	0.079	04:39
Air Blank	0.000	04:39
Diagnostics Check	OK	04:39

Data Download Date/Time: 12/10/2018 20:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEADOWS,GEOFFREY,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CROFT,BRETT
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 11/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A3
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: MECCARIELLO

DATE/TIME OF TEST: 12/08/2018 03:49:02

Diagnostics Check	OK	03:41
Air Blank	0.000	03:41
Control Test	0.080	03:41
Air Blank	0.000	03:42
Subject Sample # 1	0.090	03:43
Breath Volume	1.167	03:43
Air Blank	0.000	03:44
Air Blank	0.000	03:46
Subject Sample # 2	0.090	03:47
Breath Volume	1.238	03:47
Air Blank	0.000	03:48
Control Test	0.080	03:48
Air Blank	0.000	03:48
Diagnostics Check	OK	03:49

Data Download Date/Time: 12/10/2018 20:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEADOWS,GEOFFREY,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CROFT,BRETT
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 11/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A3
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:13

SUBJECT'S LAST NAME: SCARBOROUGH

DATE/TIME OF TEST: 12/08/2018 19:39:50

Diagnostics Check	OK	19:33
Air Blank	0.000	19:33
Control Test	0.081	19:33
Air Blank	0.000	19:34
Subject Sample # 1	0.251	19:35
Breath Volume	2.457	19:35
Air Blank	0.000	19:35
Air Blank	0.000	19:37
Subject Sample # 2	0.246	19:38
Breath Volume	2.441	19:38
Air Blank	0.000	19:38
Control Test	0.081	19:39
Air Blank	0.000	19:39
Diagnostics Check	OK	19:39

Data Download Date/Time: 12/10/2018 20:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEADOWS,GEOFFREY,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CROFT,BRETT
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 11/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A3
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: SPEAKMAN

DATE/TIME OF TEST: 12/09/2018 01:29:22

Diagnostics Check	OK	01:24	
Air Blank	0.000	01:24	
Control Test	0.081	01:25	
Air Blank	0.000	01:25	
Subject Sample	# 1	01:27	Subject Test Refused
Breath Volume	0.000	01:27	
Air Blank	0.000	01:28	
Control Test	0.081	01:28	
Air Blank	0.000	01:29	
Diagnostics Check	OK	01:29	

Data Download Date/Time: 12/10/2018 20:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO,I,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: RITCHIE,FRANCIS
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 12/10/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 34917080A3
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 12/12/2018 01:57:04

Diagnostics Check	OK	01:49
Air Blank	0.000	01:50
Control Test	0.079	01:50
Air Blank	0.000	01:50
Subject Sample # 1	0.147	01:52
Breath Volume	2.070	01:52
Air Blank	0.000	01:53
Air Blank	0.000	01:55
Subject Sample # 2	0.144	01:55
Breath Volume	1.648	01:55
Air Blank	0.000	01:56
Control Test	0.079	01:56
Air Blank	0.000	01:56
Diagnostics Check	OK	01:57

Data Download Date/Time: 01/09/2019 19:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,PATRICK,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CROFT,BRETT
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 12/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A3
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: MADISON

DATE/TIME OF TEST: 12/12/2018 23:09:10

Diagnostics Check	OK	23:02
Air Blank	0.000	23:03
Control Test	0.081	23:03
Air Blank	0.000	23:04
Subject Sample # 1	0.090	23:04
Breath Volume	1.949	23:04
Air Blank	0.000	23:05
Air Blank	0.000	23:07
Subject Sample # 2	0.091	23:07
Breath Volume	2.011	23:07
Air Blank	0.000	23:08
Control Test	0.080	23:08
Air Blank	0.000	23:09
Diagnostics Check	OK	23:09

Data Download Date/Time: 01/09/2019 19:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,PATRICK,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CROFT,BRETT
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 12/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A3
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: AHLEFEID

DATE/TIME OF TEST: 12/14/2018 03:22:55

Diagnostics Check	OK	03:14
Air Blank	0.000	03:14
Control Test	0.080	03:15
Air Blank	0.000	03:15
Subject Sample # 1	0.227	03:18
Breath Volume	1.265	03:18
Air Blank	0.000	03:19
Air Blank	0.000	03:20
Subject Sample # 2	0.227	03:21
Breath Volume	1.167	03:21
Air Blank	0.000	03:22
Control Test	0.077	03:22
Air Blank	0.000	03:22
Diagnostics Check	OK	03:22

Data Download Date/Time: 01/09/2019 19:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO,IVAN,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: RITCHIE,FRANCIS
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 12/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A3
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 12/16/2018 01:45:55

Diagnostics Check	OK	01:42	
Air Blank	0.000	01:43	
Control Test	0.080	01:43	
Air Blank	0.000	01:44	
Subject Sample	# 1	01:44	Subject Test Refused
Breath Volume	0.000	01:44	
Air Blank	0.000	01:45	
Control Test	0.081	01:45	
Air Blank	0.000	01:45	
Diagnostics Check	OK	01:45	

Data Download Date/Time: 01/09/2019 19:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEADOWS,GEOFFREY,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: VARVARO,JESSE
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 12/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A3
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: LITAKER

DATE/TIME OF TEST: 12/18/2018 19:26:31

Diagnostics Check	OK	19:20
Air Blank	0.000	19:20
Control Test	0.081	19:20
Air Blank	0.000	19:21
Subject Sample # 1	0.035	19:22
Breath Volume	2.363	19:22
Air Blank	0.000	19:22
Air Blank	0.000	19:24
Subject Sample # 2	0.032	19:25
Breath Volume	2.046	19:25
Air Blank	0.000	19:25
Control Test	0.080	19:25
Air Blank	0.000	19:26
Diagnostics Check	OK	19:26

Data Download Date/Time: 01/09/2019 19:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEADOWS,GEOFFREY,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CROFT,BRETT
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 12/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A3
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: BUTTERFIELD

DATE/TIME OF TEST: 12/19/2018 00:26:48

Diagnostics Check	OK	00:20
Air Blank	0.000	00:20
Control Test	0.081	00:21
Air Blank	0.000	00:21
Subject Sample # 1	0.044	00:22
Breath Volume	1.632	00:22
Air Blank	0.000	00:22
Air Blank	0.000	00:24
Subject Sample # 2	0.044	00:25
Breath Volume	2.121	00:25
Air Blank	0.000	00:25
Control Test	0.080	00:26
Air Blank	0.000	00:26
Diagnostics Check	OK	00:26

Data Download Date/Time: 01/09/2019 19:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO,IVAN,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: SIMMONS,MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/10/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 34917080A3
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 12/20/2018 00:17:28

Diagnostics Check	OK	00:14	
Air Blank	0.000	00:14	
Control Test	0.080	00:15	
Air Blank	0.000	00:15	
Subject Sample	# 1	00:16	Subject Test Refused
Breath Volume	0.000	00:16	
Air Blank	0.000	00:16	
Control Test	0.080	00:16	
Air Blank	0.000	00:17	
Diagnostics Check	OK	00:17	

Data Download Date/Time: 01/09/2019 19:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO,IVAN,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: RITCHIE,FRANCIS
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 12/10/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 34917080A3
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: GARNER

DATE/TIME OF TEST: 12/20/2018 02:00:33

Diagnostics Check	OK	01:54
Air Blank	0.000	01:54
Control Test	0.080	01:54
Air Blank	0.000	01:55
Subject Sample # 1	0.205	01:55
Breath Volume	2.148	01:55
Air Blank	0.000	01:56
Air Blank	0.000	01:58
Subject Sample # 2	0.208	01:58
Breath Volume	2.812	01:58
Air Blank	0.000	01:59
Control Test	0.078	01:59
Air Blank	0.000	02:00
Diagnostics Check	OK	02:00

Data Download Date/Time: 01/09/2019 19:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO,IVAN,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: RITCHIE,FRANCIS
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 12/10/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 34917080A3
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: HOWIE

DATE/TIME OF TEST: 12/21/2018 01:23:22

Diagnostics Check	OK	01:20	
Air Blank	0.000	01:20	
Control Test	0.080	01:21	
Air Blank	0.000	01:21	
Subject Sample	# 1	01:22	Subject Test Refused
Breath Volume	0.000	01:22	
Air Blank	0.000	01:22	
Control Test	0.081	01:22	
Air Blank	0.000	01:23	
Diagnostics Check	OK	01:23	

Data Download Date/Time: 01/09/2019 19:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,PATRICK,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CROFT,BRETT
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 12/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A3
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: ZEUNER

DATE/TIME OF TEST: 12/23/2018 00:16:12

Diagnostics Check	OK	00:09
Air Blank	0.000	00:10
Control Test	0.081	00:10
Air Blank	0.000	00:10
Subject Sample # 1	0.107	00:11
Breath Volume	1.628	00:11
Air Blank	0.000	00:12
Air Blank	0.000	00:14
Subject Sample # 2	0.103	00:14
Breath Volume	1.417	00:14
Air Blank	0.000	00:15
Control Test	0.080	00:15
Air Blank	0.000	00:16
Diagnostics Check	OK	00:16

Data Download Date/Time: 01/09/2019 19:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEADOWS,GEOFFREY,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CROFT,BRETT
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 12/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A3
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:59

SUBJECT'S LAST NAME: SCORZA

DATE/TIME OF TEST: 12/23/2018 05:25:46

Diagnostics Check	OK	05:19
Air Blank	0.000	05:19
Control Test	0.081	05:19
Air Blank	0.000	05:20
Subject Sample # 1	0.122	05:21
Breath Volume	2.175	05:21
Air Blank	0.000	05:21
Air Blank	0.000	05:23
Subject Sample # 2	0.124	05:24
Breath Volume	2.511	05:24
Air Blank	0.000	05:24
Control Test	0.079	05:25
Air Blank	0.000	05:25
Diagnostics Check	OK	05:25

Data Download Date/Time: 01/09/2019 19:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RITCHIE,FRANCIS,V
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: RITCHIE,FRANCIS
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 12/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A3
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 10:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: BLOOM

DATE/TIME OF TEST: 12/24/2018 23:27:40

Diagnostics Check	OK	23:20
Air Blank	0.000	23:21
Control Test	0.080	23:21
Air Blank	0.000	23:22
Subject Sample # 1	0.149	23:23
Breath Volume	1.718	23:23
Air Blank	0.000	23:23
Air Blank	0.000	23:25
Subject Sample # 2	0.157	23:26
Breath Volume	1.550	23:26
Air Blank	0.000	23:26
Control Test	0.080	23:27
Air Blank	0.000	23:27
Diagnostics Check	OK	23:27

Data Download Date/Time: 01/09/2019 19:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RITCHIE,FRANCIS,V
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: RITCHIE,FRANCIS
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 12/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A3
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: FOSS

DATE/TIME OF TEST: 12/25/2018 02:37:40

Diagnostics Check	OK	02:30
Air Blank	0.000	02:30
Control Test	0.079	02:30
Air Blank	0.000	02:31
Subject Sample # 1	0.097	02:32
Breath Volume	3.910	02:32
Air Blank	0.000	02:33
Air Blank	0.000	02:34
Subject Sample # 2	0.090	02:36
Breath Volume	1.867	02:36
Air Blank	0.000	02:36
Control Test	0.079	02:37
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

Data Download Date/Time: 01/09/2019 19:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KJEKSTAD,ERIC,C
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HELM,DEREK
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A3
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: DAMBROSIO

DATE/TIME OF TEST: 12/27/2018 01:00:54

Diagnostics Check	OK	00:53
Air Blank	0.000	00:53
Control Test	0.080	00:53
Air Blank	0.000	00:54
Subject Sample # 1	0.124	00:56
Breath Volume	2.027	00:56
Air Blank	0.000	00:57
Air Blank	0.000	00:58
Subject Sample # 2	0.120	00:59
Breath Volume	1.746	00:59
Air Blank	0.000	00:59
Control Test	0.079	01:00
Air Blank	0.000	01:00
Diagnostics Check	OK	01:00

Data Download Date/Time: 01/09/2019 19:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROFT,BRETT,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CROFT,BRETT
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 12/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A3
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 12/27/2018 01:34:20

Diagnostics Check	OK	01:30	
Air Blank	0.000	01:31	
Control Test	0.080	01:31	
Air Blank	0.000	01:31	
Subject Sample	# 1	01:32	Subject Test Refused
Breath Volume	0.000	01:32	
Air Blank	0.000	01:33	
Control Test	0.079	01:33	
Air Blank	0.000	01:34	
Diagnostics Check	OK	01:34	

Data Download Date/Time: 01/09/2019 19:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROFT,BRETT,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CROFT,BRETT
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 12/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A3
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:29

SUBJECT'S LAST NAME: TIMNEY

DATE/TIME OF TEST: 12/27/2018 20:56:31

Diagnostics Check	OK	20:49
Air Blank	0.000	20:50
Control Test	0.080	20:50
Air Blank	0.000	20:51
Subject Sample # 1	0.000	20:52
Breath Volume	1.511	20:52
Air Blank	0.000	20:52
Air Blank	0.000	20:54
Subject Sample # 2	0.000	20:55
Breath Volume	1.539	20:55
Air Blank	0.000	20:55
Control Test	0.081	20:55
Air Blank	0.000	20:56
Diagnostics Check	OK	20:56

Data Download Date/Time: 01/09/2019 19:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELLS,AUSTIN,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KIBBEY,BRADLEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A3
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 12/29/2018 03:54:33

Diagnostics Check	OK	03:47
Air Blank	0.000	03:48
Control Test	0.081	03:48
Air Blank	0.000	03:49
Subject Sample # 1	0.187	03:49
Breath Volume	2.343	03:49
Air Blank	0.000	03:50
Air Blank	0.000	03:52
Subject Sample # 2	0.188	03:52
Breath Volume	2.011	03:52
Air Blank	0.000	03:53
Control Test	0.080	03:53
Air Blank	0.000	03:54
Diagnostics Check	OK	03:54

Data Download Date/Time: 01/09/2019 19:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RITCHIE, FRANCIS, V
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: JOHNSON, J
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 12/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A3
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 12/30/2018 01:30:48

Diagnostics Check	OK	01:24
Air Blank	0.000	01:24
Control Test	0.081	01:24
Air Blank	0.000	01:25
Subject Sample # 1	0.111	01:26
Breath Volume	3.023	01:26
Air Blank	0.000	01:26
Air Blank	0.000	01:28
Subject Sample # 2	0.109	01:29
Breath Volume	3.152	01:29
Air Blank	0.000	01:29
Control Test	0.080	01:30
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

Data Download Date/Time: 01/09/2019 19:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RITCHIE,FRANCIS,V
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: RITCHIE,FRANCIS
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 12/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A3
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 12:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: CUTTING

DATE/TIME OF TEST: 12/30/2018 02:12:51

Diagnostics Check	OK	02:01	
Air Blank	0.000	02:02	
Control Test	0.080	02:02	
Air Blank	0.000	02:02	
Subject Sample	# 1 0.191	02:06	
Breath Volume	2.308	02:06	
Air Blank	0.000	02:06	
Air Blank	0.000	02:08	
Subject Sample	# 2	02:11	Subject Test Refused
Breath Volume	0.000	02:11	
Air Blank	0.000	02:11	
Control Test	0.079	02:12	
Air Blank	0.000	02:12	
Diagnostics Check	OK	02:12	

Data Download Date/Time: 01/09/2019 19:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RITCHIE,FRANCIS,V
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: RITCHIE,FRANCIS
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 12/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A3
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: ROOK

DATE/TIME OF TEST: 12/31/2018 00:51:20

Diagnostics Check	OK	00:44
Air Blank	0.000	00:45
Control Test	0.081	00:45
Air Blank	0.000	00:45
Subject Sample # 1	0.238	00:46
Breath Volume	2.320	00:46
Air Blank	0.000	00:47
Air Blank	0.000	00:49
Subject Sample # 2	0.236	00:49
Breath Volume	2.683	00:49
Air Blank	0.000	00:50
Control Test	0.080	00:50
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

Data Download Date/Time: 01/09/2019 19:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROFT,BRETT,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CROFT,BRETT
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 12/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A3
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: DORTA

DATE/TIME OF TEST: 12/31/2018 23:36:54

Diagnostics Check	OK	23:32	
Air Blank	0.000	23:33	
Control Test	0.081	23:33	
Air Blank	0.000	23:33	
Subject Sample	# 1	23:35	Subject Test Refused
Breath Volume	0.000	23:35	
Air Blank	0.000	23:36	
Control Test	0.081	23:36	
Air Blank	0.000	23:36	
Diagnostics Check	OK	23:36	

Data Download Date/Time: 01/09/2019 19:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICOTTILLI,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI,JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: MONTGOMERY

DATE/TIME OF TEST: 12/01/2018 01:40:44

Diagnostics Check	OK	01:37	
Air Blank	0.000	01:38	
Control Test	0.082	01:38	
Air Blank	0.000	01:38	
Subject Sample	# 1	01:39	Subject Test Refused
Breath Volume	0.000	01:39	
Air Blank	0.000	01:39	
Control Test	0.082	01:40	
Air Blank	0.000	01:40	
Diagnostics Check	OK	01:40	

Data Download Date/Time: 12/26/2018 21:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: SANDERS

DATE/TIME OF TEST: 12/01/2018 04:23:16

Diagnostics Check	OK	04:17
Air Blank	0.000	04:17
Control Test	0.082	04:17
Air Blank	0.000	04:18
Subject Sample # 1	0.103	04:18
Breath Volume	1.511	04:18
Air Blank	0.000	04:19
Air Blank	0.000	04:21
Subject Sample # 2	0.102	04:21
Breath Volume	1.687	04:21
Air Blank	0.000	04:22
Control Test	0.081	04:22
Air Blank	0.000	04:23
Diagnostics Check	OK	04:23

Data Download Date/Time: 12/26/2018 21:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICOTTILLI,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI,JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 04:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:54

SUBJECT'S LAST NAME: KRAPFEL

DATE/TIME OF TEST: 12/01/2018 05:27:08

Diagnostics Check	OK	05:14	
Air Blank	0.000	05:14	
Control Test	0.081	05:15	
Air Blank	0.000	05:15	
Subject Sample	# 1	05:18	No Sample Provided
Breath Volume	0.000	05:18	
Air Blank	0.000	05:19	
Air Blank	0.000	05:21	
Subject Sample	# 2	0.182	05:22
Breath Volume	1.859	05:22	
Air Blank	0.000	05:22	
Air Blank	0.000	05:24	
Subject Sample	# 3	05:25	Subject Test Refused
Breath Volume	0.000	05:25	
Air Blank	0.000	05:26	
Control Test	0.080	05:26	
Air Blank	0.000	05:27	
Diagnostics Check	OK	05:27	

Data Download Date/Time: 12/26/2018 21:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNAVELY,KENT,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SNAVELY,KENT
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 16:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 12/01/2018 17:16:50

Diagnostics Check	OK	17:10
Air Blank	0.000	17:10
Control Test	0.081	17:10
Air Blank	0.000	17:11
Subject Sample # 1	0.143	17:12
Breath Volume	3.480	17:12
Air Blank	0.000	17:12
Air Blank	0.000	17:14
Subject Sample # 2	0.142	17:15
Breath Volume	3.335	17:15
Air Blank	0.000	17:15
Control Test	0.081	17:16
Air Blank	0.000	17:16
Diagnostics Check	OK	17:16

Data Download Date/Time: 12/26/2018 21:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 12/02/2018 04:24:46

Diagnostics Check	OK	04:18
Air Blank	0.000	04:19
Control Test	0.082	04:19
Air Blank	0.000	04:19
Subject Sample # 1	0.099	04:20
Breath Volume	1.589	04:20
Air Blank	0.000	04:20
Air Blank	0.000	04:22
Subject Sample # 2	0.098	04:23
Breath Volume	1.542	04:23
Air Blank	0.000	04:23
Control Test	0.081	04:24
Air Blank	0.000	04:24
Diagnostics Check	OK	04:24

Data Download Date/Time: 12/26/2018 21:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 20:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:56

SUBJECT'S LAST NAME: SHAPLAND

DATE/TIME OF TEST: 12/05/2018 21:22:58

Diagnostics Check	OK	21:16
Air Blank	0.000	21:17
Control Test	0.082	21:17
Air Blank	0.000	21:18
Subject Sample # 1	0.313	21:18
Breath Volume	2.828	21:18
Air Blank	0.000	21:19
Air Blank	0.000	21:20
Subject Sample # 2	0.323	21:21
Breath Volume	3.332	21:21
Air Blank	0.000	21:22
Control Test	0.082	21:22
Air Blank	0.000	21:22
Diagnostics Check	OK	21:22

Data Download Date/Time: 12/26/2018 21:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICOTTILLI, JOSHUA, A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI, JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 12/06/2018 02:08:46

Diagnostics Check	OK	02:04	
Air Blank	0.000	02:04	
Control Test	0.083	02:05	
Air Blank	0.000	02:05	
Subject Sample	# 1	02:07	Subject Test Refused
Breath Volume	0.000	02:07	
Air Blank	0.000	02:07	
Control Test	0.083	02:08	
Air Blank	0.000	02:08	
Diagnostics Check	OK	02:08	

Data Download Date/Time: 12/26/2018 21:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERSAUD,COMPTON,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PERSAUD,COMPTON
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 11/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 19:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:34

SUBJECT'S LAST NAME: MCAULEY

DATE/TIME OF TEST: 12/06/2018 20:05:45

Diagnostics Check	OK	19:54	
Air Blank	0.000	19:54	
Control Test	0.082	19:55	
Air Blank	0.000	19:55	
Subject Sample	# 1	19:58	No Sample Provided
Breath Volume	0.000	19:58	
Air Blank	0.000	19:59	
Air Blank	0.000	20:01	
Subject Sample	# 2	20:04	No Sample Provided
Breath Volume	0.000	20:04	
Air Blank	0.000	20:04	
Control Test	0.082	20:05	
Air Blank	0.000	20:05	
Diagnostics Check	OK	20:05	

Data Download Date/Time: 12/26/2018 21:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KAHVEDZIC, SEMIR,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BEHRENSHOUSER, RICHARD
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 07:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 12/06/2018 21:52:48

Diagnostics Check	OK	21:45
Air Blank	0.000	21:46
Control Test	0.082	21:46
Air Blank	0.000	21:47
Subject Sample # 1	0.165	21:47
Breath Volume	1.769	21:47
Air Blank	0.000	21:48
Air Blank	0.000	21:50
Subject Sample # 2	0.159	21:51
Breath Volume	2.937	21:51
Air Blank	0.000	21:51
Control Test	0.082	21:52
Air Blank	0.000	21:52
Diagnostics Check	OK	21:52

Data Download Date/Time: 12/26/2018 21:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 22:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: ALBERTSON

DATE/TIME OF TEST: 12/06/2018 23:13:17

Diagnostics Check	OK	23:06
Air Blank	0.000	23:06
Control Test	0.082	23:07
Air Blank	0.000	23:07
Subject Sample # 1	0.144	23:08
Breath Volume	1.792	23:08
Air Blank	0.000	23:09
Air Blank	0.000	23:11
Subject Sample # 2	0.139	23:11
Breath Volume	1.589	23:11
Air Blank	0.000	23:12
Control Test	0.082	23:12
Air Blank	0.000	23:13
Diagnostics Check	OK	23:13

Data Download Date/Time: 12/26/2018 21:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICOTTILLI, JOSHUA, A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI, JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 12/07/2018 05:02:12

Diagnostics Check	OK	04:53
Air Blank	0.000	04:53
Control Test	0.083	04:53
Air Blank	0.000	04:54
Subject Sample # 1	0.140	04:57
Breath Volume	1.671	04:57
Air Blank	0.000	04:57
Air Blank	0.000	04:59
Subject Sample # 2	0.142	05:00
Breath Volume	1.691	05:00
Air Blank	0.000	05:01
Control Test	0.082	05:01
Air Blank	0.000	05:02
Diagnostics Check	OK	05:02

Data Download Date/Time: 12/26/2018 21:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 12/07/2018 23:54:19

Diagnostics Check	OK	23:51	
Air Blank	0.000	23:51	
Control Test	0.082	23:52	
Air Blank	0.000	23:52	
Subject Sample	# 1	23:52	Subject Test Refused
Breath Volume	0.000	23:52	
Air Blank	0.000	23:53	
Control Test	0.082	23:53	
Air Blank	0.000	23:54	
Diagnostics Check	OK	23:54	

Data Download Date/Time: 12/26/2018 21:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICOTTILLI, JOSHUA, A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI, JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 12/08/2018 01:29:04

Diagnostics Check	OK	01:21
Air Blank	0.000	01:22
Control Test	0.082	01:22
Air Blank	0.000	01:23
Subject Sample # 1	0.000	01:24
Breath Volume	1.496	01:24
Air Blank	0.000	01:25
Air Blank	0.000	01:27
Subject Sample # 2	0.000	01:27
Breath Volume	1.742	01:27
Air Blank	0.000	01:28
Control Test	0.080	01:28
Air Blank	0.000	01:28
Diagnostics Check	OK	01:29

Data Download Date/Time: 12/26/2018 21:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: RUDKIN

DATE/TIME OF TEST: 12/08/2018 03:07:59

Diagnostics Check	OK	02:56	
Air Blank	0.000	02:57	
Control Test	0.082	02:57	
Air Blank	0.000	02:57	
Subject Sample	# 1	03:01	No Sample Provided
Breath Volume	0.000	03:01	
Air Blank	0.000	03:01	
Air Blank	0.000	03:03	
Subject Sample	# 2	03:06	No Sample Provided
Breath Volume	0.000	03:06	
Air Blank	0.000	03:07	
Control Test	0.082	03:07	
Air Blank	0.000	03:07	
Diagnostics Check	OK	03:07	

Data Download Date/Time: 12/26/2018 21:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICOTTILLI,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI,JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 05:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:54

SUBJECT'S LAST NAME: LEMACHER

DATE/TIME OF TEST: 12/08/2018 06:22:47

Diagnostics Check	OK	06:15
Air Blank	0.000	06:15
Control Test	0.081	06:16
Air Blank	0.000	06:16
Subject Sample # 1	0.194	06:17
Breath Volume	1.566	06:17
Air Blank	0.000	06:18
Air Blank	0.000	06:20
Subject Sample # 2	0.199	06:21
Breath Volume	1.847	06:21
Air Blank	0.000	06:21
Control Test	0.081	06:22
Air Blank	0.000	06:22
Diagnostics Check	OK	06:22

Data Download Date/Time: 12/26/2018 21:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE,DEPUTY WILLIAM,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: COOK,DEPUTY C.
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 18:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:56

SUBJECT'S LAST NAME: HARDY

DATE/TIME OF TEST: 12/11/2018 19:26:03

Diagnostics Check	OK	19:17	
Air Blank	0.000	19:18	
Control Test	0.082	19:18	
Air Blank	0.000	19:18	
Subject Sample	# 1	19:22	No Sample Provided
Breath Volume	0.000	19:22	
Air Blank	0.000	19:22	
Air Blank	0.000	19:24	
Subject Sample	# 2	19:24	Subject Test Refused
Breath Volume	0.000	19:24	
Air Blank	0.000	19:25	
Control Test	0.082	19:25	
Air Blank	0.000	19:25	
Diagnostics Check	OK	19:26	

Data Download Date/Time: 12/26/2018 21:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: SEMLITZ

DATE/TIME OF TEST: 12/13/2018 02:16:08

Diagnostics Check	OK	02:09
Air Blank	0.000	02:10
Control Test	0.082	02:10
Air Blank	0.000	02:11
Subject Sample # 1	0.000	02:11
Breath Volume	1.753	02:11
Air Blank	0.000	02:12
Air Blank	0.000	02:14
Subject Sample # 2	0.000	02:14
Breath Volume	2.285	02:14
Air Blank	0.000	02:15
Control Test	0.081	02:15
Air Blank	0.000	02:16
Diagnostics Check	OK	02:16

Data Download Date/Time: 12/26/2018 21:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICOTTILLI, JOSHUA, A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI, JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: PARCEL

DATE/TIME OF TEST: 12/13/2018 03:15:37

Diagnostics Check	OK	03:08
Air Blank	0.000	03:09
Control Test	0.082	03:09
Air Blank	0.000	03:10
Subject Sample # 1	0.047	03:11
Breath Volume	2.117	03:11
Air Blank	0.000	03:11
Air Blank	0.000	03:13
Subject Sample # 2	0.045	03:14
Breath Volume	1.703	03:14
Air Blank	0.000	03:14
Control Test	0.081	03:15
Air Blank	0.000	03:15
Diagnostics Check	OK	03:15

Data Download Date/Time: 12/26/2018 21:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEDE,BRENNEN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WEDE,BRENNEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: HAUK

DATE/TIME OF TEST: 12/14/2018 01:54:21

Diagnostics Check	OK	01:50	
Air Blank	0.000	01:51	
Control Test	0.082	01:51	
Air Blank	0.000	01:52	
Subject Sample	# 1	01:52	Subject Test Refused
Breath Volume	0.000	01:52	
Air Blank	0.000	01:53	
Control Test	0.082	01:53	
Air Blank	0.000	01:54	
Diagnostics Check	OK	01:54	

Data Download Date/Time: 12/26/2018 21:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICOTTILLI,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI,JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 04:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:54

SUBJECT'S LAST NAME: ROHLOFF

DATE/TIME OF TEST: 12/14/2018 05:27:20

Diagnostics Check	OK	05:14
Air Blank	0.000	05:15
Control Test	0.082	05:15
Air Blank	0.000	05:16
Subject Sample # 1	0.256	05:18
Breath Volume	1.476	05:18
Air Blank	0.000	05:19
Air Blank	0.000	05:21
Subject Sample # 2	0.227	05:22
Breath Volume	1.207	05:22
Air Blank	0.000	05:23
Air Blank	0.000	05:24
Subject Sample # 3	0.215	05:25
Breath Volume	1.324	05:25
Air Blank	0.000	05:26
Control Test	0.081	05:26
Air Blank	0.000	05:27
Diagnostics Check	OK	05:27

Data Download Date/Time: 12/26/2018 21:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: RUSH

DATE/TIME OF TEST: 12/15/2018 01:02:29

Diagnostics Check	OK	00:55
Air Blank	0.000	00:55
Control Test	0.082	00:55
Air Blank	0.000	00:56
Subject Sample # 1	0.166	00:58
Breath Volume	1.121	00:58
Air Blank	0.000	00:58
Air Blank	0.000	01:00
Subject Sample # 2	0.173	01:00
Breath Volume	2.183	01:00
Air Blank	0.000	01:01
Control Test	0.082	01:01
Air Blank	0.000	01:02
Diagnostics Check	OK	01:02

Data Download Date/Time: 12/26/2018 21:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MCKENZIE,DEPUTY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 12:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:00

SUBJECT'S LAST NAME: HOGAN

DATE/TIME OF TEST: 12/15/2018 13:23:36

Diagnostics Check	OK	13:20	
Air Blank	0.000	13:20	
Control Test	0.081	13:21	
Air Blank	0.000	13:21	
Subject Sample	# 1	13:22	Subject Test Refused
Breath Volume	0.000	13:22	
Air Blank	0.000	13:22	
Control Test	0.081	13:23	
Air Blank	0.000	13:23	
Diagnostics Check	OK	13:23	

Data Download Date/Time: 12/26/2018 21:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, WILLIAM, T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: FRENCH

DATE/TIME OF TEST: 12/15/2018 21:24:05

Diagnostics Check	OK	21:17
Air Blank	0.000	21:17
Control Test	0.082	21:18
Air Blank	0.000	21:18
Subject Sample # 1	0.173	21:19
Breath Volume	1.644	21:19
Air Blank	0.000	21:20
Air Blank	0.000	21:21
Subject Sample # 2	0.170	21:22
Breath Volume	1.644	21:22
Air Blank	0.000	21:23
Control Test	0.082	21:23
Air Blank	0.000	21:24
Diagnostics Check	OK	21:24

Data Download Date/Time: 12/26/2018 21:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: REDDON

DATE/TIME OF TEST: 12/16/2018 04:49:13

Diagnostics Check	OK	04:43
Air Blank	0.000	04:43
Control Test	0.082	04:43
Air Blank	0.000	04:44
Subject Sample # 1	0.056	04:44
Breath Volume	2.574	04:44
Air Blank	0.000	04:45
Air Blank	0.000	04:47
Subject Sample # 2	0.049	04:47
Breath Volume	1.175	04:47
Air Blank	0.000	04:48
Control Test	0.082	04:48
Air Blank	0.000	04:49
Diagnostics Check	OK	04:49

Data Download Date/Time: 12/26/2018 21:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE,DEPUTY WILLIAM,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WILTSE,DEPUTY WILLIAM
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 04:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: ARCE

DATE/TIME OF TEST: 12/16/2018 05:43:53

Diagnostics Check	OK	05:40	
Air Blank	0.000	05:41	
Control Test	0.082	05:41	
Air Blank	0.000	05:42	
Subject Sample	# 1	05:42	Subject Test Refused
Breath Volume	0.000	05:42	
Air Blank	0.000	05:42	
Control Test	0.082	05:43	
Air Blank	0.000	05:43	
Diagnostics Check	OK	05:43	

Data Download Date/Time: 12/26/2018 21:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE,DEPUTY WILLIAM,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: VAN WORP,DEPUTY J.
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 05:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:35

SUBJECT'S LAST NAME: LYONS

DATE/TIME OF TEST: 12/16/2018 06:04:17

Diagnostics Check	OK	05:57
Air Blank	0.000	05:57
Control Test	0.081	05:57
Air Blank	0.000	05:58
Subject Sample # 1	0.215	05:59
Breath Volume	2.132	05:59
Air Blank	0.000	05:59
Air Blank	0.000	06:01
Subject Sample # 2	0.207	06:02
Breath Volume	1.671	06:02
Air Blank	0.000	06:03
Control Test	0.080	06:03
Air Blank	0.000	06:04
Diagnostics Check	OK	06:04

Data Download Date/Time: 12/26/2018 21:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOHR,KRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOHR,K
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 19:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: UMBERTO

DATE/TIME OF TEST: 12/16/2018 20:29:03

Diagnostics Check	OK	20:26	
Air Blank	0.000	20:26	
Control Test	0.081	20:26	
Air Blank	0.000	20:27	
Subject Sample	# 1	20:27	RFI Detect
Breath Volume	0.000	20:27	
Air Blank	0.000	20:28	
Control Test	0.081	20:28	
Air Blank	0.000	20:28	
Diagnostics Check	OK	20:29	

Data Download Date/Time: 12/26/2018 21:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOHR,KRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOHR,KRISTOPHER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 19:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: UMBERTO

DATE/TIME OF TEST: 12/16/2018 20:36:07

Diagnostics Check	OK	20:31	
Air Blank	0.000	20:32	
Control Test	0.082	20:32	
Air Blank	0.000	20:32	
Subject Sample	# 1	20:35	RFI Detect
Breath Volume	0.019	20:35	
Air Blank		20:35	RFI Detect
Air Blank		20:36	RFI Detect

Data Download Date/Time: 12/26/2018 21:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE,DEPUTY WILLIAM,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CAMPION,DEPUTY K.
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: MCGUIRE

DATE/TIME OF TEST: 12/17/2018 03:36:16

Diagnostics Check	OK	03:29
Air Blank	0.000	03:29
Control Test	0.082	03:30
Air Blank	0.000	03:30
Subject Sample # 1	0.186	03:31
Breath Volume	2.605	03:31
Air Blank	0.000	03:32
Air Blank	0.000	03:34
Subject Sample # 2	0.167	03:34
Breath Volume	1.683	03:34
Air Blank	0.000	03:35
Control Test	0.081	03:35
Air Blank	0.000	03:36
Diagnostics Check	OK	03:36

Data Download Date/Time: 12/26/2018 21:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUDBRINK,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SUDBRINK,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 22:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: TURMAN

DATE/TIME OF TEST: 12/17/2018 22:54:41

Diagnostics Check	OK	22:48
Air Blank	0.000	22:49
Control Test	0.082	22:49
Air Blank	0.000	22:49
Subject Sample # 1	0.096	22:50
Breath Volume	2.371	22:50
Air Blank	0.000	22:50
Air Blank	0.000	22:52
Subject Sample # 2	0.089	22:53
Breath Volume	2.601	22:53
Air Blank	0.000	22:53
Control Test	0.081	22:54
Air Blank	0.000	22:54
Diagnostics Check	OK	22:54

Data Download Date/Time: 12/26/2018 21:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: MAYES

DATE/TIME OF TEST: 12/18/2018 22:47:45

Diagnostics Check	OK	22:41
Air Blank	0.000	22:41
Control Test	0.082	22:42
Air Blank	0.000	22:42
Subject Sample # 1	0.212	22:43
Breath Volume	2.566	22:43
Air Blank	0.000	22:44
Air Blank	0.000	22:45
Subject Sample # 2	0.207	22:46
Breath Volume	2.218	22:46
Air Blank	0.000	22:46
Control Test	0.082	22:47
Air Blank	0.000	22:47
Diagnostics Check	OK	22:47

Data Download Date/Time: 12/26/2018 21:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: SHEPHERD

DATE/TIME OF TEST: 12/19/2018 02:56:55

Diagnostics Check	OK	02:50
Air Blank	0.000	02:50
Control Test	0.083	02:51
Air Blank	0.000	02:51
Subject Sample # 1	0.125	02:52
Breath Volume	1.511	02:52
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample # 2	0.123	02:55
Breath Volume	1.406	02:55
Air Blank	0.000	02:56
Control Test	0.082	02:56
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 12/26/2018 21:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEDE,BRENNEN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WEDE,BRENNEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: AMOS

DATE/TIME OF TEST: 12/19/2018 04:11:27

Diagnostics Check	OK	04:05
Air Blank	0.000	04:05
Control Test	0.081	04:05
Air Blank	0.000	04:06
Subject Sample # 1	0.217	04:06
Breath Volume	2.250	04:06
Air Blank	0.000	04:07
Air Blank	0.000	04:09
Subject Sample # 2	0.217	04:09
Breath Volume	2.566	04:09
Air Blank	0.000	04:10
Control Test	0.082	04:10
Air Blank	0.000	04:11
Diagnostics Check	OK	04:11

Data Download Date/Time: 12/26/2018 21:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICOTTILLI, JOSHUA, A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI, JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 12/20/2018 00:27:07

Diagnostics Check	OK	00:15	
Air Blank	0.000	00:16	
Control Test	0.081	00:16	
Air Blank	0.000	00:17	
Subject Sample	# 1	00:20	No Sample Provided
Breath Volume	0.000	00:20	
Air Blank	0.000	00:20	
Air Blank	0.000	00:22	
Subject Sample	# 2	00:25	No Sample Provided
Breath Volume	0.000	00:25	
Air Blank	0.000	00:26	
Control Test	0.082	00:26	
Air Blank	0.000	00:27	
Diagnostics Check	OK	00:27	

Data Download Date/Time: 12/26/2018 21:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: HILLIER

DATE/TIME OF TEST: 12/20/2018 04:15:08

Diagnostics Check	OK	04:11	
Air Blank	0.000	04:11	
Control Test	0.080	04:11	
Air Blank	0.000	04:12	
Subject Sample	# 1	04:13	Subject Test Refused
Breath Volume	0.000	04:13	
Air Blank	0.000	04:14	
Control Test	0.082	04:14	
Air Blank	0.000	04:15	
Diagnostics Check	OK	04:15	

Data Download Date/Time: 12/26/2018 21:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOODNIGHT,MICHAEL,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GOODNIGHT,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: ZACKERU

DATE/TIME OF TEST: 12/21/2018 00:57:06

Diagnostics Check	OK	00:51
Air Blank	0.000	00:51
Control Test	0.082	00:51
Air Blank	0.000	00:52
Subject Sample # 1	0.185	00:52
Breath Volume	1.332	00:52
Air Blank	0.000	00:53
Air Blank	0.000	00:55
Subject Sample # 2	0.192	00:55
Breath Volume	1.964	00:55
Air Blank	0.000	00:56
Control Test	0.082	00:56
Air Blank	0.000	00:57
Diagnostics Check	OK	00:57

Data Download Date/Time: 12/26/2018 21:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: COWLES

DATE/TIME OF TEST: 12/21/2018 01:38:45

Diagnostics Check	OK	01:32
Air Blank	0.000	01:33
Control Test	0.081	01:33
Air Blank	0.000	01:34
Subject Sample # 1	0.000	01:34
Breath Volume	1.480	01:34
Air Blank	0.000	01:34
Air Blank	0.000	01:36
Subject Sample # 2	0.000	01:37
Breath Volume	1.531	01:37
Air Blank	0.000	01:37
Control Test	0.082	01:38
Air Blank	0.000	01:38
Diagnostics Check	OK	01:38

Data Download Date/Time: 12/26/2018 21:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICOTTILLI, JOSHUA, A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI, JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 04:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:01

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 12/21/2018 05:30:46

Diagnostics Check	OK	05:21
Air Blank	0.000	05:21
Control Test	0.082	05:22
Air Blank	0.000	05:22
Subject Sample # 1	0.243	05:25
Breath Volume	1.500	05:25
Air Blank	0.000	05:26
Air Blank	0.000	05:27
Subject Sample # 2	0.237	05:29
Breath Volume	1.839	05:29
Air Blank	0.000	05:29
Control Test	0.079	05:30
Air Blank	0.000	05:30
Diagnostics Check	OK	05:30

Data Download Date/Time: 12/26/2018 21:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUDBRINK,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SUDBRINK,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:28

SUBJECT'S LAST NAME: TODD

DATE/TIME OF TEST: 12/22/2018 21:54:34

Diagnostics Check	OK	21:48
Air Blank	0.000	21:49
Control Test	0.082	21:49
Air Blank	0.000	21:49
Subject Sample # 1	0.174	21:50
Breath Volume	1.878	21:50
Air Blank	0.000	21:50
Air Blank	0.000	21:52
Subject Sample # 2	0.171	21:53
Breath Volume	1.414	21:53
Air Blank	0.000	21:53
Control Test	0.081	21:53
Air Blank	0.000	21:54
Diagnostics Check	OK	21:54

Data Download Date/Time: 12/26/2018 21:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE,DEPUTY WILLIAM,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: COOK,DEPUTY C.
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: LOMBARDO 3

DATE/TIME OF TEST: 12/25/2018 01:43:57

Diagnostics Check	OK	01:40	
Air Blank	0.000	01:41	
Control Test	0.083	01:41	
Air Blank	0.000	01:42	
Subject Sample	# 1	01:42	Subject Test Refused
Breath Volume	0.000	01:42	
Air Blank	0.000	01:43	
Control Test	0.082	01:43	
Air Blank	0.000	01:43	
Diagnostics Check	OK	01:43	

Data Download Date/Time: 12/26/2018 21:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE,DEPUTY WILLIAM,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HOBART,OFFICER C.
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: DYER

DATE/TIME OF TEST: 12/25/2018 05:38:47

Diagnostics Check	OK	05:31
Air Blank	0.000	05:31
Control Test	0.082	05:32
Air Blank	0.000	05:32
Subject Sample # 1	0.288	05:33
Breath Volume	1.390	05:33
Air Blank	0.000	05:34
Air Blank	0.000	05:36
Subject Sample # 2	0.302	05:37
Breath Volume	2.292	05:37
Air Blank	0.000	05:37
Control Test	0.081	05:38
Air Blank	0.000	05:38
Diagnostics Check	OK	05:38

Data Download Date/Time: 12/26/2018 21:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEYER, CHARLES, L
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: GEYER, C
ARRESTING OFFICER AGENCY: Perry PD
LAST AGENCY INSPECTION: 11/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: OTHER
ARREST TIME: 10:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:55

SUBJECT'S LAST NAME: BOLES

DATE/TIME OF TEST: 12/03/2018 11:22:19

Diagnostics Check	OK	11:15
Air Blank	0.000	11:16
Control Test	0.080	11:16
Air Blank	0.000	11:17
Subject Sample # 1	0.000	11:17
Breath Volume	3.386	11:17
Air Blank	0.000	11:18
Air Blank	0.000	11:20
Subject Sample # 2	0.000	11:20
Breath Volume	3.492	11:20
Air Blank	0.000	11:21
Control Test	0.079	11:21
Air Blank	0.000	11:22
Diagnostics Check	OK	11:22

Data Download Date/Time: 01/02/2019 18:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BASS,JASON,P
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: BASS,JASON
ARRESTING OFFICER AGENCY: Perry PD
LAST AGENCY INSPECTION: 11/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: OTHER
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: BOOKER

DATE/TIME OF TEST: 12/08/2018 01:46:20

Diagnostics Check	OK	01:43	
Air Blank	0.000	01:43	
Control Test	0.080	01:43	
Air Blank	0.000	01:44	
Subject Sample	# 1	01:45	Sequence Aborted
Breath Volume	0.000	01:45	
Air Blank	0.000	01:46	

Data Download Date/Time: 01/02/2019 18:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL, CHARLES, J
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: MITCHELL, CHARLES
ARRESTING OFFICER AGENCY: Perry PD
LAST AGENCY INSPECTION: 11/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: RECTOR

DATE/TIME OF TEST: 12/15/2018 02:52:21

Diagnostics Check	OK	02:46
Air Blank	0.000	02:46
Control Test	0.080	02:46
Air Blank	0.000	02:47
Subject Sample # 1	0.141	02:47
Breath Volume	1.734	02:47
Air Blank	0.000	02:48
Air Blank	0.000	02:50
Subject Sample # 2	0.136	02:50
Breath Volume	2.660	02:50
Air Blank	0.000	02:51
Control Test	0.080	02:51
Air Blank	0.000	02:52
Diagnostics Check	OK	02:52

Data Download Date/Time: 01/02/2019 18:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KURTZ,THEODORE,F
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: ALEVISATOS,JARED
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 609878
DRY GAS EXPIRY DATE: 11/01/2019
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: BLANCHARD

DATE/TIME OF TEST: 12/04/2018 00:56:32

Diagnostics Check	OK	00:50
Air Blank	0.000	00:50
Control Test	0.079	00:51
Air Blank	0.000	00:51
Subject Sample # 1	0.222	00:52
Breath Volume	1.542	00:52
Air Blank	0.000	00:52
Air Blank	0.000	00:54
Subject Sample # 2	0.215	00:55
Breath Volume	1.781	00:55
Air Blank	0.000	00:55
Control Test	0.079	00:55
Air Blank	0.000	00:56
Diagnostics Check	OK	00:56

Data Download Date/Time: 12/31/2018 16:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KURTZ,THEODORE,F
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: SCHNIBBEN,BRIAN
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 609878
DRY GAS EXPIRY DATE: 11/01/2019
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: PAULEY

DATE/TIME OF TEST: 12/05/2018 02:52:20

Diagnostics Check	OK	02:45
Air Blank	0.000	02:46
Control Test	0.080	02:46
Air Blank	0.000	02:47
Subject Sample # 1	0.182	02:47
Breath Volume	1.531	02:47
Air Blank	0.000	02:48
Air Blank	0.000	02:50
Subject Sample # 2	0.196	02:50
Breath Volume	1.761	02:50
Air Blank	0.000	02:51
Control Test	0.079	02:51
Air Blank	0.000	02:52
Diagnostics Check	OK	02:52

Data Download Date/Time: 12/31/2018 16:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KURTZ,THEODORE,F
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: SCHNIBBEN,BRIAN
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 609878
DRY GAS EXPIRY DATE: 11/01/2019
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: ELLISON

DATE/TIME OF TEST: 12/09/2018 03:40:08

Diagnostics Check	OK	03:34
Air Blank	0.000	03:34
Control Test	0.079	03:34
Air Blank	0.000	03:35
Subject Sample # 1	0.077	03:35
Breath Volume	2.175	03:35
Air Blank	0.000	03:36
Air Blank	0.000	03:38
Subject Sample # 2	0.073	03:38
Breath Volume	1.398	03:38
Air Blank	0.000	03:39
Control Test	0.079	03:39
Air Blank	0.000	03:40
Diagnostics Check	OK	03:40

Data Download Date/Time: 12/31/2018 16:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KURTZ,THEODORE,F
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: PORTILLO,SILVIO
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 609878
DRY GAS EXPIRY DATE: 11/01/2019
VIOLATION CODE: DUI
ARREST TIME: 04:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 12/09/2018 04:54:40

Diagnostics Check	OK	04:47
Air Blank	0.000	04:48
Control Test	0.080	04:48
Air Blank	0.000	04:49
Subject Sample # 1	0.131	04:49
Breath Volume	1.246	04:49
Air Blank	0.000	04:50
Air Blank	0.000	04:52
Subject Sample # 2	0.135	04:53
Breath Volume	1.484	04:53
Air Blank	0.000	04:53
Control Test	0.079	04:54
Air Blank	0.000	04:54
Diagnostics Check	OK	04:54

Data Download Date/Time: 12/31/2018 16:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDWARDS,JUSTIN,D
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: HAMBSH,C
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 609878
DRY GAS EXPIRY DATE: 11/01/2019
VIOLATION CODE: DUI
ARREST TIME: 10:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:45

SUBJECT'S LAST NAME: MAUSOLF

DATE/TIME OF TEST: 12/23/2018 11:15:59

Diagnostics Check	OK	11:08
Air Blank	0.000	11:08
Control Test	0.081	11:09
Air Blank	0.000	11:09
Subject Sample # 1	0.047	11:11
Breath Volume	1.160	11:11
Air Blank	0.000	11:11
Air Blank	0.000	11:13
Subject Sample # 2	0.043	11:14
Breath Volume	1.183	11:14
Air Blank	0.000	11:15
Control Test	0.079	11:15
Air Blank	0.000	11:15
Diagnostics Check	OK	11:15

Data Download Date/Time: 12/31/2018 16:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOLAN,JONATHAN,W
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: GUILLERM,MICHAEL
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 609878
DRY GAS EXPIRY DATE: 11/01/2019
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 12/24/2018 02:14:22

Diagnostics Check	OK	02:07
Air Blank	0.000	02:07
Control Test	0.079	02:07
Air Blank	0.000	02:08
Subject Sample # 1	0.203	02:09
Breath Volume	2.042	02:09
Air Blank	0.000	02:09
Air Blank	0.000	02:11
Subject Sample # 2	0.211	02:12
Breath Volume	2.156	02:12
Air Blank	0.000	02:13
Control Test	0.078	02:13
Air Blank	0.000	02:14
Diagnostics Check	OK	02:14

Data Download Date/Time: 12/31/2018 16:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDWARDS,JUSTIN,D
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: HAMBSH,C
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 609878
DRY GAS EXPIRY DATE: 11/01/2019
VIOLATION CODE: DUI
ARREST TIME: 08:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:50

SUBJECT'S LAST NAME: PEARSON

DATE/TIME OF TEST: 12/27/2018 10:19:25

Diagnostics Check	OK	10:13
Air Blank	0.000	10:13
Control Test	0.080	10:14
Air Blank	0.000	10:14
Subject Sample # 1	0.000	10:15
Breath Volume	3.207	10:15
Air Blank	0.000	10:15
Air Blank	0.000	10:17
Subject Sample # 2	0.000	10:18
Breath Volume	2.648	10:18
Air Blank	0.000	10:18
Control Test	0.081	10:18
Air Blank	0.000	10:19
Diagnostics Check	OK	10:19

Data Download Date/Time: 12/31/2018 16:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,L,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: WINCELOWICZ,L
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 11/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: PEPLINSKI

DATE/TIME OF TEST: 12/02/2018 02:45:53

Diagnostics Check	OK	02:37
Air Blank	0.000	02:38
Control Test	0.082	02:38
Air Blank	0.000	02:39
Subject Sample # 1	0.077	02:40
Breath Volume	2.320	02:40
Air Blank	0.000	02:40
Air Blank	0.000	02:42
Subject Sample # 2	0.076	02:44
Breath Volume	2.265	02:44
Air Blank	0.000	02:44
Control Test	0.083	02:45
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 12/12/2018 13:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCBRIDE,TIMOTHY,A
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: MCBRIDE,TIMOTHY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 11/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: FUTCH

DATE/TIME OF TEST: 12/02/2018 03:12:10

Diagnostics Check	OK	03:05
Air Blank	0.000	03:06
Control Test	0.082	03:06
Air Blank	0.000	03:07
Subject Sample # 1	0.000	03:07
Breath Volume	1.988	03:07
Air Blank	0.000	03:08
Air Blank	0.000	03:10
Subject Sample # 2	0.000	03:10
Breath Volume	1.750	03:10
Air Blank	0.000	03:11
Control Test	0.082	03:11
Air Blank	0.000	03:12
Diagnostics Check	OK	03:12

Data Download Date/Time: 12/12/2018 13:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOTELLO,A,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: BOTELLO,A
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 11/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:56

SUBJECT'S LAST NAME: DEBOY

DATE/TIME OF TEST: 12/02/2018 21:31:55

Diagnostics Check	OK	21:25
Air Blank	0.000	21:25
Control Test	0.082	21:26
Air Blank	0.000	21:26
Subject Sample # 1	0.132	21:27
Breath Volume	3.636	21:27
Air Blank	0.000	21:28
Air Blank	0.000	21:29
Subject Sample # 2	0.128	21:30
Breath Volume	3.945	21:30
Air Blank	0.000	21:30
Control Test	0.081	21:31
Air Blank	0.000	21:31
Diagnostics Check	OK	21:31

Data Download Date/Time: 12/12/2018 13:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCBRIDE,TIMOTHY,A
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: MCBRIDE,TIMOTHY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 11/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: RUSSELL

DATE/TIME OF TEST: 12/03/2018 01:41:20

Diagnostics Check	OK	01:34
Air Blank	0.000	01:34
Control Test	0.082	01:35
Air Blank	0.000	01:35
Subject Sample # 1	0.176	01:36
Breath Volume	1.804	01:36
Air Blank	0.000	01:37
Air Blank	0.000	01:38
Subject Sample # 2	0.173	01:39
Breath Volume	1.855	01:39
Air Blank	0.000	01:40
Control Test	0.081	01:40
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

Data Download Date/Time: 12/12/2018 13:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,EARTHEN,A
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: BROWN,EARTHEN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 11/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: PERDUE

DATE/TIME OF TEST: 12/04/2018 03:17:04

Diagnostics Check	OK	03:10
Air Blank	0.000	03:10
Control Test	0.082	03:10
Air Blank	0.000	03:11
Subject Sample # 1	0.231	03:11
Breath Volume	1.468	03:11
Air Blank	0.000	03:12
Air Blank	0.000	03:14
Subject Sample # 2	0.227	03:15
Breath Volume	1.664	03:15
Air Blank	0.000	03:16
Control Test	0.080	03:16
Air Blank	0.000	03:16
Diagnostics Check	OK	03:17

Data Download Date/Time: 12/12/2018 13:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEINER,RYAN,T
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: WALSH,L
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 11/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 12/07/2018 03:47:33

Diagnostics Check	OK	03:35	
Air Blank	0.000	03:35	
Control Test	0.082	03:36	
Air Blank	0.000	03:36	
Subject Sample	# 1 0.199	03:39	Volume Not Met
Breath Volume	0.875	03:39	
Air Blank	0.000	03:40	
Air Blank	0.000	03:42	
Subject Sample	# 2 0.214	03:42	
Breath Volume	1.234	03:42	
Air Blank	0.000	03:43	
Air Blank	0.000	03:45	
Subject Sample	# 3 0.221	03:45	
Breath Volume	1.832	03:45	
Air Blank	0.000	03:46	
Control Test	0.079	03:46	
Air Blank	0.000	03:47	
Diagnostics Check	OK	03:47	

Data Download Date/Time: 12/12/2018 13:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,EARTHEN,A
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: BROWN,EARTHEN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 11/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 12/08/2018 03:29:35

Diagnostics Check	OK	03:22
Air Blank	0.000	03:23
Control Test	0.083	03:23
Air Blank	0.000	03:23
Subject Sample # 1	0.132	03:24
Breath Volume	2.347	03:24
Air Blank	0.000	03:25
Air Blank	0.000	03:27
Subject Sample # 2	0.121	03:27
Breath Volume	2.363	03:27
Air Blank	0.000	03:28
Control Test	0.080	03:28
Air Blank	0.000	03:29
Diagnostics Check	OK	03:29

Data Download Date/Time: 12/12/2018 13:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINTON, ADAM, H
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: WINTON, ADAM
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 11/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: BUI
ARREST TIME: 17:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:43

SUBJECT'S LAST NAME: WESSON

DATE/TIME OF TEST: 12/08/2018 20:10:45

Diagnostics Check	OK	20:04
Air Blank	0.000	20:04
Control Test	0.082	20:04
Air Blank	0.000	20:05
Subject Sample # 1	0.125	20:05
Breath Volume	3.507	20:05
Air Blank	0.000	20:06
Air Blank	0.000	20:08
Subject Sample # 2	0.123	20:08
Breath Volume	3.945	20:08
Air Blank	0.000	20:09
Control Test	0.081	20:10
Air Blank	0.000	20:10
Diagnostics Check	OK	20:10

Data Download Date/Time: 12/12/2018 13:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,JAMES,R
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ADAMS,JAMES
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 11/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 12/11/2018 01:23:54

Diagnostics Check	OK	01:16
Air Blank	0.000	01:16
Control Test	0.083	01:16
Air Blank	0.000	01:17
Subject Sample # 1	0.182	01:18
Breath Volume	2.015	01:18
Air Blank	0.000	01:19
Air Blank	0.000	01:21
Subject Sample # 2	0.180	01:22
Breath Volume	2.160	01:22
Air Blank	0.000	01:22
Control Test	0.081	01:23
Air Blank	0.000	01:23
Diagnostics Check	OK	01:23

Data Download Date/Time: 12/12/2018 13:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSH,DANIEL,C
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: RUSH,DANIEL
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: RHIEL

DATE/TIME OF TEST: 12/13/2018 04:25:54

Diagnostics Check	OK	04:19
Air Blank	0.000	04:19
Control Test	0.083	04:19
Air Blank	0.000	04:20
Subject Sample # 1	0.200	04:21
Breath Volume	2.648	04:21
Air Blank	0.000	04:22
Air Blank	0.000	04:23
Subject Sample # 2	0.199	04:24
Breath Volume	2.910	04:24
Air Blank	0.000	04:24
Control Test	0.079	04:25
Air Blank	0.000	04:25
Diagnostics Check	OK	04:25

Data Download Date/Time: 01/09/2019 15:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,EARTHEN,A
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SHOEMAKER,JACLYN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: VAIL

DATE/TIME OF TEST: 12/14/2018 03:55:43

Diagnostics Check	OK	03:48
Air Blank	0.000	03:49
Control Test	0.083	03:49
Air Blank	0.000	03:49
Subject Sample # 1	0.149	03:50
Breath Volume	1.468	03:50
Air Blank	0.000	03:51
Air Blank	0.000	03:53
Subject Sample # 2	0.147	03:53
Breath Volume	1.285	03:53
Air Blank	0.000	03:54
Control Test	0.080	03:55
Air Blank	0.000	03:55
Diagnostics Check	OK	03:55

Data Download Date/Time: 01/09/2019 15:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROMAN,A,C
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GREGGORY,T
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:01

SUBJECT'S LAST NAME: RYAN

DATE/TIME OF TEST: 12/14/2018 16:31:29

Diagnostics Check	OK	16:21
Air Blank	0.000	16:21
Control Test	0.081	16:22
Air Blank	0.000	16:22
Subject Sample # 1	0.072	16:24
Breath Volume	1.164	16:24
Air Blank	0.000	16:24
Air Blank	0.000	16:26
Subject Sample # 2	0.083	16:29
Breath Volume	2.144	16:29
Air Blank	0.000	16:30
Control Test	0.082	16:30
Air Blank	0.000	16:31
Diagnostics Check	OK	16:31

Data Download Date/Time: 01/09/2019 15:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSEPH,G
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: RYAN,MATTHEW
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: THORNTON

DATE/TIME OF TEST: 12/15/2018 00:30:05

Diagnostics Check	OK	00:22
Air Blank	0.000	00:22
Control Test	0.082	00:23
Air Blank	0.000	00:23
Subject Sample # 1	0.000	00:25
Breath Volume	1.546	00:25
Air Blank	0.000	00:25
Air Blank	0.000	00:27
Subject Sample # 2	0.000	00:28
Breath Volume	2.019	00:28
Air Blank	0.000	00:29
Control Test	0.081	00:29
Air Blank	0.000	00:30
Diagnostics Check	OK	00:30

Data Download Date/Time: 01/09/2019 15:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,CHRISTINE,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: HALL,CHRISTINE
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:56

SUBJECT'S LAST NAME: SEXTON

DATE/TIME OF TEST: 12/15/2018 19:23:42

Diagnostics Check	OK	19:17
Air Blank	0.000	19:17
Control Test	0.083	19:17
Air Blank	0.000	19:18
Subject Sample # 1	0.136	19:19
Breath Volume	3.316	19:19
Air Blank	0.000	19:19
Air Blank	0.000	19:21
Subject Sample # 2	0.135	19:21
Breath Volume	3.800	19:21
Air Blank	0.000	19:22
Control Test	0.082	19:23
Air Blank	0.000	19:23
Diagnostics Check	OK	19:23

Data Download Date/Time: 01/09/2019 15:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,JAMES,R
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ADAMS,JAMES
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: CROWLEY

DATE/TIME OF TEST: 12/15/2018 21:37:13

Diagnostics Check	OK	21:28
Air Blank	0.000	21:28
Control Test	0.082	21:28
Air Blank	0.000	21:29
Subject Sample # 1	0.242	21:32
Breath Volume	2.773	21:32
Air Blank	0.000	21:33
Air Blank	0.000	21:34
Subject Sample # 2	0.257	21:35
Breath Volume	2.453	21:35
Air Blank	0.000	21:36
Control Test	0.078	21:36
Air Blank	0.000	21:37
Diagnostics Check	OK	21:37

Data Download Date/Time: 01/09/2019 15:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHAEFER,KENDALL,A
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: SCHAEFER,KENDALL
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 12/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: PERZANOWSKI

DATE/TIME OF TEST: 12/16/2018 02:28:24

Diagnostics Check	OK	02:22
Air Blank	0.000	02:22
Control Test	0.080	02:22
Air Blank	0.000	02:23
Subject Sample # 1	0.207	02:23
Breath Volume	1.492	02:23
Air Blank	0.000	02:24
Air Blank	0.000	02:26
Subject Sample # 2	0.213	02:26
Breath Volume	1.109	02:26
Air Blank	0.000	02:27
Control Test	0.081	02:27
Air Blank	0.000	02:28
Diagnostics Check	OK	02:28

Data Download Date/Time: 01/09/2019 15:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEINER, RYAN, T
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: MORGAN, R
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 12/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: PEREZ LOPEZ

DATE/TIME OF TEST: 12/16/2018 04:21:10

Diagnostics Check	OK	04:12
Air Blank	0.000	04:12
Control Test	0.082	04:12
Air Blank	0.000	04:13
Subject Sample # 1	0.146	04:16
Breath Volume	1.457	04:16
Air Blank	0.000	04:17
Air Blank	0.000	04:18
Subject Sample # 2	0.136	04:19
Breath Volume	1.117	04:19
Air Blank	0.000	04:20
Control Test	0.079	04:20
Air Blank	0.000	04:21
Diagnostics Check	OK	04:21

Data Download Date/Time: 01/09/2019 15:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROMAN,A,C
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: DAVILA,M
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 12/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 08:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:30

SUBJECT'S LAST NAME: PIERROT

DATE/TIME OF TEST: 12/16/2018 09:00:55

Diagnostics Check	OK	08:50
Air Blank	0.000	08:51
Control Test	0.082	08:51
Air Blank	0.000	08:52
Subject Sample # 1	0.191	08:52
Breath Volume	2.304	08:52
Air Blank	0.000	08:53
Air Blank	0.000	08:55
Subject Sample # 2	0.170	08:56
Breath Volume	1.316	08:56
Air Blank	0.000	08:57
Air Blank	0.000	08:58
Subject Sample # 3	0.176	08:59
Breath Volume	1.207	08:59
Air Blank	0.000	08:59
Control Test	0.080	09:00
Air Blank	0.000	09:00
Diagnostics Check	OK	09:00

Data Download Date/Time: 01/09/2019 15:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEINER,RYAN,T
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: LEON,B
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 12/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: BERNARDO AMAYA

DATE/TIME OF TEST: 12/17/2018 04:53:13

Diagnostics Check	OK	04:46
Air Blank	0.000	04:46
Control Test	0.083	04:46
Air Blank	0.000	04:47
Subject Sample # 1	0.201	04:48
Breath Volume	1.656	04:48
Air Blank	0.000	04:49
Air Blank	0.000	04:50
Subject Sample # 2	0.205	04:51
Breath Volume	1.476	04:51
Air Blank	0.000	04:52
Control Test	0.081	04:52
Air Blank	0.000	04:53
Diagnostics Check	OK	04:53

Data Download Date/Time: 01/09/2019 15:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROMAN,A,C
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: PIERSON,C
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 10:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:17

SUBJECT'S LAST NAME: RAY

DATE/TIME OF TEST: 12/19/2018 11:42:01

Diagnostics Check	OK	11:37	
Air Blank	0.000	11:38	
Control Test	0.081	11:38	
Air Blank	0.000	11:39	
Subject Sample # 1	0.313	11:40	
Breath Volume	1.582	11:40	
Air Blank		11:41	Purge Fail
Air Blank	0.000	11:42	

Data Download Date/Time: 01/09/2019 15:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROMAN,A,C
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: PIERSON,C
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 10:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:17

SUBJECT'S LAST NAME: RAY

DATE/TIME OF TEST: 12/19/2018 11:50:16

Diagnostics Check	OK	11:43
Air Blank	0.000	11:43
Control Test	0.078	11:43
Air Blank	0.000	11:44
Subject Sample # 1	0.346	11:44
Breath Volume	1.742	11:44
Air Blank	0.000	11:45
Air Blank	0.000	11:47
Subject Sample # 2	0.348	11:48
Breath Volume	2.156	11:48
Air Blank	0.000	11:49
Control Test	0.078	11:49
Air Blank	0.000	11:50
Diagnostics Check	OK	11:50

Data Download Date/Time: 01/09/2019 15:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERRANO,ROMAN',
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SERRANO,ROMAN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:58

SUBJECT'S LAST NAME: BONENBERGER

DATE/TIME OF TEST: 12/19/2018 20:33:49

Diagnostics Check	OK	20:25
Air Blank	0.000	20:25
Control Test	0.081	20:26
Air Blank	0.000	20:26
Subject Sample # 1	0.000	20:27
Breath Volume	1.300	20:27
Air Blank	0.000	20:28
Air Blank	0.000	20:30
Subject Sample # 2	0.000	20:32
Breath Volume	1.257	20:32
Air Blank	0.000	20:32
Control Test	0.081	20:33
Air Blank	0.000	20:33
Diagnostics Check	OK	20:33

Data Download Date/Time: 01/09/2019 15:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDIN,JEREMY,L
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: ARTHUR,MELVIN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: VITALIANO

DATE/TIME OF TEST: 12/22/2018 05:51:14

Diagnostics Check	OK	05:44
Air Blank	0.000	05:44
Control Test	0.082	05:44
Air Blank	0.000	05:45
Subject Sample # 1	0.111	05:46
Breath Volume	1.878	05:46
Air Blank	0.000	05:46
Air Blank	0.000	05:48
Subject Sample # 2	0.101	05:49
Breath Volume	1.425	05:49
Air Blank	0.000	05:50
Control Test	0.079	05:50
Air Blank	0.000	05:51
Diagnostics Check	OK	05:51

Data Download Date/Time: 01/09/2019 15:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERONI, KRISTI,
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: NERONI, KRISTI
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 12/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: SCHEID

DATE/TIME OF TEST: 12/23/2018 04:25:54

Diagnostics Check	OK	04:19
Air Blank	0.000	04:19
Control Test	0.082	04:19
Air Blank	0.000	04:20
Subject Sample # 1	0.138	04:21
Breath Volume	2.214	04:21
Air Blank	0.000	04:21
Air Blank	0.000	04:23
Subject Sample # 2	0.135	04:24
Breath Volume	2.203	04:24
Air Blank	0.000	04:24
Control Test	0.082	04:25
Air Blank	0.000	04:25
Diagnostics Check	OK	04:25

Data Download Date/Time: 01/09/2019 15:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDIN,JEREMY,L
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GARRETT,RYAN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 08:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:07

SUBJECT'S LAST NAME: JOVEL AMAYA

DATE/TIME OF TEST: 12/23/2018 10:36:00

Diagnostics Check	OK	10:29
Air Blank	0.000	10:29
Control Test	0.080	10:30
Air Blank	0.000	10:30
Subject Sample # 1	0.096	10:31
Breath Volume	1.375	10:31
Air Blank	0.000	10:31
Air Blank	0.000	10:33
Subject Sample # 2	0.093	10:34
Breath Volume	1.652	10:34
Air Blank	0.000	10:35
Control Test	0.078	10:35
Air Blank	0.000	10:35
Diagnostics Check	OK	10:36

Data Download Date/Time: 01/09/2019 15:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,EARTHEN,A
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: HIDEG,J
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: ESPOSITO

DATE/TIME OF TEST: 12/24/2018 00:27:13

Diagnostics Check	OK	00:20
Air Blank	0.000	00:20
Control Test	0.082	00:21
Air Blank	0.000	00:21
Subject Sample # 1	0.236	00:22
Breath Volume	1.605	00:22
Air Blank	0.000	00:23
Air Blank	0.000	00:24
Subject Sample # 2	0.229	00:25
Breath Volume	2.039	00:25
Air Blank	0.000	00:26
Control Test	0.079	00:26
Air Blank	0.000	00:27
Diagnostics Check	OK	00:27

Data Download Date/Time: 01/09/2019 15:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,EARTHEN,A
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: ROSE,C
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: MEDLEY

DATE/TIME OF TEST: 12/24/2018 03:11:29

Diagnostics Check	OK	03:04
Air Blank	0.000	03:05
Control Test	0.082	03:05
Air Blank	0.000	03:06
Subject Sample # 1	0.025	03:06
Breath Volume	3.105	03:06
Air Blank	0.000	03:07
Air Blank	0.000	03:09
Subject Sample # 2	0.025	03:09
Breath Volume	2.863	03:09
Air Blank	0.000	03:10
Control Test	0.081	03:10
Air Blank	0.000	03:11
Diagnostics Check	OK	03:11

Data Download Date/Time: 01/09/2019 15:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,EARTHEN,A
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: MELIDOR,B
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 12/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 12/27/2018 04:39:33

Diagnostics Check	OK	04:31
Air Blank	0.000	04:32
Control Test	0.081	04:32
Air Blank	0.000	04:33
Subject Sample # 1	0.190	04:34
Breath Volume	2.921	04:34
Air Blank	0.000	04:35
Air Blank	0.000	04:37
Subject Sample # 2	0.186	04:37
Breath Volume	3.132	04:37
Air Blank	0.000	04:38
Control Test	0.078	04:38
Air Blank	0.000	04:39
Diagnostics Check	OK	04:39

Data Download Date/Time: 01/09/2019 15:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHAEFER, KENDALL, A
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: CAIL, M
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 13:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: PICKARD

DATE/TIME OF TEST: 12/27/2018 15:28:16

Diagnostics Check	OK	15:21
Air Blank	0.000	15:21
Control Test	0.081	15:22
Air Blank	0.000	15:22
Subject Sample # 1	0.266	15:23
Breath Volume	1.660	15:23
Air Blank	0.000	15:23
Air Blank	0.000	15:25
Subject Sample # 2	0.250	15:26
Breath Volume	1.496	15:26
Air Blank	0.000	15:27
Control Test	0.079	15:27
Air Blank	0.000	15:28
Diagnostics Check	OK	15:28

Data Download Date/Time: 01/09/2019 15:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WITT,KELLY,
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: WITT,KELLY
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 12/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: JIMENEZ

DATE/TIME OF TEST: 12/29/2018 02:31:49

Diagnostics Check	OK	02:24	
Air Blank	0.000	02:25	
Control Test	0.081	02:25	
Air Blank	0.000	02:26	
Subject Sample # 1	0.342	02:26	
Breath Volume	2.000	02:26	
Air Blank	0.000	02:27	
Air Blank	0.000	02:29	
Subject Sample # 2	0.340	02:30	
Breath Volume	1.679	02:30	
Air Blank		02:31	Purge Fail
Air Blank	0.000	02:31	

Data Download Date/Time: 01/09/2019 15:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINTON, ADAM, H
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: WINTON, ADAM
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 12/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: BUI
ARREST TIME: 18:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:57

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 12/29/2018 20:25:06

Diagnostics Check	OK	20:18
Air Blank	0.000	20:18
Control Test	0.081	20:18
Air Blank	0.000	20:19
Subject Sample # 1	0.200	20:19
Breath Volume	1.371	20:19
Air Blank	0.000	20:20
Air Blank	0.000	20:22
Subject Sample # 2	0.194	20:23
Breath Volume	2.828	20:23
Air Blank	0.000	20:24
Control Test	0.080	20:24
Air Blank	0.000	20:25
Diagnostics Check	OK	20:25

Data Download Date/Time: 01/09/2019 15:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JOSIAH,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SMITH,JOSIAH
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: PICALLO HERNANDEZ

DATE/TIME OF TEST: 12/30/2018 02:35:35

Diagnostics Check	OK	02:27
Air Blank	0.000	02:28
Control Test	0.081	02:28
Air Blank	0.000	02:29
Subject Sample # 1	0.192	02:30
Breath Volume	2.070	02:30
Air Blank	0.000	02:31
Air Blank	0.000	02:33
Subject Sample # 2	0.188	02:33
Breath Volume	1.562	02:33
Air Blank	0.000	02:34
Control Test	0.079	02:34
Air Blank	0.000	02:35
Diagnostics Check	OK	02:35

Data Download Date/Time: 01/09/2019 15:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,JAMES,R
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ADAMS,JAMES
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: KOENIG

DATE/TIME OF TEST: 12/31/2018 04:43:19

Diagnostics Check	OK	04:35
Air Blank	0.000	04:36
Control Test	0.082	04:36
Air Blank	0.000	04:36
Subject Sample # 1	0.249	04:38
Breath Volume	1.476	04:38
Air Blank	0.000	04:39
Air Blank	0.000	04:40
Subject Sample # 2	0.258	04:41
Breath Volume	1.640	04:41
Air Blank	0.000	04:42
Control Test	0.078	04:42
Air Blank	0.000	04:43
Diagnostics Check	OK	04:43

Data Download Date/Time: 01/09/2019 15:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAETSCH,JOSH,E
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: LEADINGHAM,ZACHARY
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: WHIGMAN

DATE/TIME OF TEST: 12/08/2018 02:59:45

Diagnostics Check	OK	02:52
Air Blank	0.000	02:52
Control Test	0.080	02:53
Air Blank	0.000	02:53
Subject Sample # 1	0.094	02:54
Breath Volume	2.683	02:54
Air Blank	0.000	02:55
Air Blank	0.000	02:57
Subject Sample # 2	0.081	02:57
Breath Volume	1.789	02:57
Air Blank	0.000	02:58
Control Test	0.079	02:59
Air Blank	0.000	02:59
Diagnostics Check	OK	02:59

Data Download Date/Time: 01/04/2019 12:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JASON,E
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PAETSCH,JOSHUA
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 12/24/2018 01:32:07

Diagnostics Check	OK	01:25
Air Blank	0.000	01:25
Control Test	0.080	01:26
Air Blank	0.000	01:26
Subject Sample # 1	0.176	01:27
Breath Volume	1.574	01:27
Air Blank	0.000	01:27
Air Blank	0.000	01:29
Subject Sample # 2	0.166	01:30
Breath Volume	2.199	01:30
Air Blank	0.000	01:31
Control Test	0.079	01:31
Air Blank	0.000	01:32
Diagnostics Check	OK	01:32

Data Download Date/Time: 01/04/2019 12:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,J,
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PAETSCH,J
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:12

SUBJECT'S LAST NAME: DEVANNY

DATE/TIME OF TEST: 12/29/2018 19:52:20

Diagnostics Check	OK	19:49	
Air Blank	0.000	19:49	
Control Test	0.080	19:50	
Air Blank	0.000	19:50	
Subject Sample	# 1	19:50	Subject Test Refused
Breath Volume	0.000	19:50	
Air Blank	0.000	19:51	
Control Test	0.080	19:51	
Air Blank	0.000	19:52	
Diagnostics Check	OK	19:52	

Data Download Date/Time: 01/04/2019 12:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JASON,E
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PAETSCH,JOSHUA
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: TROSSBACH

DATE/TIME OF TEST: 12/31/2018 01:00:33

Diagnostics Check	OK	00:52
Air Blank	0.000	00:53
Control Test	0.080	00:53
Air Blank	0.000	00:54
Subject Sample # 1	0.055	00:54
Breath Volume	1.695	00:54
Air Blank	0.000	00:55
Air Blank	0.000	00:57
Subject Sample # 2	0.053	00:58
Breath Volume	1.335	00:58
Air Blank	0.000	00:59
Control Test	0.079	00:59
Air Blank	0.000	01:00
Diagnostics Check	OK	01:00

Data Download Date/Time: 01/04/2019 12:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 11/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: BOOTHE

DATE/TIME OF TEST: 12/01/2018 04:50:00

Diagnostics Check	OK	04:47	
Air Blank	0.000	04:47	
Control Test	0.078	04:47	
Air Blank	0.000	04:48	
Subject Sample	# 1	04:48	Subject Test Refused
Breath Volume	0.000	04:48	
Air Blank	0.000	04:49	
Control Test	0.077	04:49	
Air Blank	0.000	04:49	
Diagnostics Check	OK	04:50	

Data Download Date/Time: 12/13/2018 10:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 11/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 05:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:34

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 12/01/2018 06:05:32

Diagnostics Check	OK	05:59
Air Blank	0.000	05:59
Control Test	0.078	05:59
Air Blank	0.000	06:00
Subject Sample # 1	0.092	06:00
Breath Volume	1.140	06:00
Air Blank	0.000	06:01
Air Blank	0.000	06:03
Subject Sample # 2	0.094	06:03
Breath Volume	2.039	06:03
Air Blank	0.000	06:04
Control Test	0.078	06:04
Air Blank	0.000	06:05
Diagnostics Check	OK	06:05

Data Download Date/Time: 12/13/2018 10:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 11/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: GONZALEZ BRAVO

DATE/TIME OF TEST: 12/02/2018 03:33:37

Diagnostics Check	OK	03:30	
Air Blank	0.000	03:31	
Control Test	0.078	03:31	
Air Blank	0.000	03:32	
Subject Sample	# 1	03:32	Subject Test Refused
Breath Volume	0.000	03:32	
Air Blank	0.000	03:32	
Control Test	0.076	03:32	
Air Blank	0.000	03:33	
Diagnostics Check	OK	03:33	

Data Download Date/Time: 12/13/2018 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: SLIMAK,MARK
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 11/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: WHEELER

DATE/TIME OF TEST: 12/05/2018 02:02:44

Diagnostics Check	OK	01:56
Air Blank	0.000	01:57
Control Test	0.077	01:57
Air Blank	0.000	01:58
Subject Sample # 1	0.148	01:58
Breath Volume	2.761	01:58
Air Blank	0.000	01:58
Air Blank	0.000	02:00
Subject Sample # 2	0.149	02:01
Breath Volume	4.015	02:01
Air Blank	0.000	02:01
Control Test	0.077	02:02
Air Blank	0.000	02:02
Diagnostics Check	OK	02:02

Data Download Date/Time: 12/13/2018 10:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 11/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: MASON

DATE/TIME OF TEST: 12/05/2018 04:27:51

Diagnostics Check	OK	04:25	
Air Blank	0.000	04:25	
Control Test	0.077	04:25	
Air Blank	0.000	04:26	
Subject Sample	# 1	04:26	Subject Test Refused
Breath Volume	0.000	04:26	
Air Blank	0.000	04:26	
Control Test	0.078	04:27	
Air Blank	0.000	04:27	
Diagnostics Check	OK	04:27	

Data Download Date/Time: 12/13/2018 10:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 11/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: CHAPARRO PRIETO

DATE/TIME OF TEST: 12/07/2018 00:54:33

Diagnostics Check	OK	00:48
Air Blank	0.000	00:48
Control Test	0.079	00:49
Air Blank	0.000	00:49
Subject Sample # 1	0.099	00:50
Breath Volume	4.453	00:50
Air Blank	0.000	00:50
Air Blank	0.000	00:52
Subject Sample # 2	0.101	00:52
Breath Volume	4.937	00:52
Air Blank	0.000	00:53
Control Test	0.078	00:53
Air Blank	0.000	00:54
Diagnostics Check	OK	00:54

Data Download Date/Time: 12/13/2018 10:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: DELGADO,MIGUEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 11/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: OCHOA

DATE/TIME OF TEST: 12/09/2018 02:37:04

Diagnostics Check	OK	02:34	
Air Blank	0.000	02:34	
Control Test	0.078	02:34	
Air Blank	0.000	02:35	
Subject Sample	# 1	02:35	Subject Test Refused
Breath Volume	0.000	02:35	
Air Blank	0.000	02:36	
Control Test	0.078	02:36	
Air Blank	0.000	02:36	
Diagnostics Check	OK	02:37	

Data Download Date/Time: 12/13/2018 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: MCAULEY,DEVARO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 11/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 05:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:36

SUBJECT'S LAST NAME: GOPALANI

DATE/TIME OF TEST: 12/09/2018 06:16:42

Diagnostics Check	OK	06:13	
Air Blank	0.000	06:14	
Control Test	0.078	06:14	
Air Blank	0.000	06:14	
Subject Sample	# 1	06:15	Subject Test Refused
Breath Volume	0.000	06:15	
Air Blank	0.000	06:15	
Control Test	0.077	06:16	
Air Blank	0.000	06:16	
Diagnostics Check	OK	06:16	

Data Download Date/Time: 12/13/2018 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 11/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: MATUTE

DATE/TIME OF TEST: 12/12/2018 04:45:23

Diagnostics Check	OK	04:38
Air Blank	0.000	04:39
Control Test	0.080	04:39
Air Blank	0.000	04:40
Subject Sample # 1	0.167	04:40
Breath Volume	2.781	04:40
Air Blank	0.000	04:41
Air Blank	0.000	04:43
Subject Sample # 2	0.165	04:43
Breath Volume	2.140	04:43
Air Blank	0.000	04:44
Control Test	0.078	04:44
Air Blank	0.000	04:45
Diagnostics Check	OK	04:45

Data Download Date/Time: 12/13/2018 10:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 11/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 05:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:39

SUBJECT'S LAST NAME: VAN UDEN

DATE/TIME OF TEST: 12/13/2018 06:17:42

Diagnostics Check	OK	06:10
Air Blank	0.000	06:10
Control Test	0.079	06:11
Air Blank	0.000	06:11
Subject Sample # 1	0.151	06:12
Breath Volume	1.812	06:12
Air Blank	0.000	06:13
Air Blank	0.000	06:15
Subject Sample # 2	0.149	06:16
Breath Volume	1.808	06:16
Air Blank	0.000	06:16
Control Test	0.078	06:17
Air Blank	0.000	06:17
Diagnostics Check	OK	06:17

Data Download Date/Time: 12/13/2018 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 11/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: WENDORFF

DATE/TIME OF TEST: 12/01/2018 04:59:54

Diagnostics Check	OK	04:53
Air Blank	0.000	04:53
Control Test	0.077	04:54
Air Blank	0.000	04:54
Subject Sample # 1	0.196	04:55
Breath Volume	1.492	04:55
Air Blank	0.000	04:56
Air Blank	0.000	04:57
Subject Sample # 2	0.193	04:58
Breath Volume	1.175	04:58
Air Blank	0.000	04:58
Control Test	0.077	04:59
Air Blank	0.000	04:59
Diagnostics Check	OK	04:59

Data Download Date/Time: 12/07/2018 09:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 11/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: HERRERA-REYES

DATE/TIME OF TEST: 12/02/2018 02:57:29

Diagnostics Check	OK	02:50
Air Blank	0.000	02:50
Control Test	0.077	02:51
Air Blank	0.000	02:51
Subject Sample # 1	0.159	02:52
Breath Volume	4.023	02:52
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample # 2	0.160	02:55
Breath Volume	3.789	02:55
Air Blank	0.000	02:56
Control Test	0.075	02:56
Air Blank	0.000	02:57
Diagnostics Check	OK	02:57

Data Download Date/Time: 12/07/2018 09:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 11/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: YORIS

DATE/TIME OF TEST: 12/05/2018 03:18:40

Diagnostics Check	OK	03:15	
Air Blank	0.000	03:16	
Control Test	0.077	03:16	
Air Blank	0.000	03:17	
Subject Sample	# 1	03:17	Subject Test Refused
Breath Volume	0.000	03:17	
Air Blank	0.000	03:17	
Control Test	0.078	03:18	
Air Blank	0.000	03:18	
Diagnostics Check	OK	03:18	

Data Download Date/Time: 12/07/2018 09:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 11/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: PEGO

DATE/TIME OF TEST: 12/07/2018 03:20:59

Diagnostics Check	OK	03:18	
Air Blank	0.000	03:18	
Control Test	0.079	03:18	
Air Blank	0.000	03:19	
Subject Sample	# 1	03:19	Subject Test Refused
Breath Volume	0.000	03:19	
Air Blank	0.000	03:20	
Control Test	0.080	03:20	
Air Blank	0.000	03:20	
Diagnostics Check	OK	03:20	

Data Download Date/Time: 12/07/2018 09:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAYERS,MICHELLE,H
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: SAYERS,MICHELLE
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 12/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: GARCES-NAVAS

DATE/TIME OF TEST: 12/15/2018 01:47:13

Diagnostics Check	OK	01:43	
Air Blank	0.000	01:43	
Control Test	0.081	01:44	
Air Blank	0.000	01:44	
Subject Sample	# 1	01:45	Subject Test Refused
Breath Volume	0.000	01:45	
Air Blank	0.000	01:46	
Control Test	0.080	01:46	
Air Blank	0.000	01:47	
Diagnostics Check	OK	01:47	

Data Download Date/Time: 01/07/2019 00:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAYERS,MICHELLE,H
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: SAYERS,MICHELLE
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 12/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: IBANEZ-FRAGA

DATE/TIME OF TEST: 12/17/2018 01:52:41

Diagnostics Check	OK	01:44
Air Blank	0.000	01:44
Control Test	0.080	01:45
Air Blank	0.000	01:45
Subject Sample # 1	0.095	01:48
Breath Volume	1.539	01:48
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample # 2	0.093	01:51
Breath Volume	1.742	01:51
Air Blank	0.000	01:51
Control Test	0.079	01:52
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

Data Download Date/Time: 01/07/2019 00:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAYERS,MICHELLE,H
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: SAYERS,MICHELLE
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 12/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: FERRER

DATE/TIME OF TEST: 12/21/2018 03:20:54

Diagnostics Check	OK	03:13
Air Blank	0.000	03:14
Control Test	0.080	03:14
Air Blank	0.000	03:15
Subject Sample # 1	0.249	03:15
Breath Volume	1.433	03:15
Air Blank	0.000	03:16
Air Blank	0.000	03:18
Subject Sample # 2	0.237	03:19
Breath Volume	1.472	03:19
Air Blank	0.000	03:19
Control Test	0.080	03:20
Air Blank	0.000	03:20
Diagnostics Check	OK	03:20

Data Download Date/Time: 01/07/2019 00:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PECCI,Z,A
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: PECCI,Z
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 11/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: BRANCH

DATE/TIME OF TEST: 12/01/2018 18:50:44

Diagnostics Check	OK	18:44
Air Blank	0.000	18:44
Control Test	0.080	18:45
Air Blank	0.000	18:45
Subject Sample # 1	0.333	18:46
Breath Volume	2.171	18:46
Air Blank	0.000	18:47
Air Blank	0.000	18:48
Subject Sample # 2	0.331	18:49
Breath Volume	1.566	18:49
Air Blank	0.000	18:49
Control Test	0.078	18:50
Air Blank	0.000	18:50
Diagnostics Check	OK	18:50

Data Download Date/Time: 12/19/2018 22:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 11/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: DEMER

DATE/TIME OF TEST: 12/02/2018 23:28:34

Diagnostics Check	OK	23:22
Air Blank	0.000	23:22
Control Test	0.079	23:22
Air Blank	0.000	23:23
Subject Sample # 1	0.184	23:23
Breath Volume	2.070	23:23
Air Blank	0.000	23:24
Air Blank	0.000	23:26
Subject Sample # 2	0.196	23:26
Breath Volume	1.476	23:26
Air Blank	0.000	23:27
Control Test	0.080	23:27
Air Blank	0.000	23:28
Diagnostics Check	OK	23:28

Data Download Date/Time: 12/19/2018 22:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREGORY,AMANDA,M
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: KASPROW,MARK
ARRESTING OFFICER AGENCY: Jupiter Island
LAST AGENCY INSPECTION: 11/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:14

SUBJECT'S LAST NAME: RITCHIE

DATE/TIME OF TEST: 12/03/2018 18:07:39

Diagnostics Check	OK	17:53	
Air Blank	0.000	17:53	
Control Test	0.080	17:54	
Air Blank	0.000	17:54	
Subject Sample	# 1 0.102	17:57	Volume Not Met
Breath Volume	0.457	17:57	
Air Blank	0.000	17:58	
Air Blank	0.000	18:00	
Subject Sample	# 2 0.129	18:01	
Breath Volume	2.804	18:01	
Air Blank	0.000	18:02	
Air Blank	0.000	18:04	
Subject Sample	# 3 0.123	18:06	
Breath Volume	1.308	18:06	
Air Blank	0.000	18:06	
Control Test	0.079	18:07	
Air Blank	0.000	18:07	
Diagnostics Check	OK	18:07	

Data Download Date/Time: 12/19/2018 22:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 11/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: THAL

DATE/TIME OF TEST: 12/08/2018 02:17:26

Diagnostics Check	OK	02:11
Air Blank	0.000	02:11
Control Test	0.081	02:11
Air Blank	0.000	02:12
Subject Sample # 1	0.160	02:12
Breath Volume	1.277	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample # 2	0.167	02:15
Breath Volume	2.183	02:15
Air Blank	0.000	02:16
Control Test	0.080	02:16
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

Data Download Date/Time: 12/19/2018 22:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 11/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: PEDRO

DATE/TIME OF TEST: 12/09/2018 04:17:09

Diagnostics Check	OK	04:10
Air Blank	0.000	04:11
Control Test	0.080	04:11
Air Blank	0.000	04:12
Subject Sample # 1	0.100	04:12
Breath Volume	2.117	04:12
Air Blank	0.000	04:13
Air Blank	0.000	04:15
Subject Sample # 2	0.097	04:15
Breath Volume	2.187	04:15
Air Blank	0.000	04:16
Control Test	0.081	04:16
Air Blank	0.000	04:17
Diagnostics Check	OK	04:17

Data Download Date/Time: 12/19/2018 22:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPLAN,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: CAPLAN,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 11/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:16

SUBJECT'S LAST NAME: SENNETT

DATE/TIME OF TEST: 12/11/2018 20:43:36

Diagnostics Check	OK	20:37
Air Blank	0.000	20:37
Control Test	0.080	20:37
Air Blank	0.000	20:38
Subject Sample # 1	0.169	20:38
Breath Volume	1.671	20:38
Air Blank	0.000	20:39
Air Blank	0.000	20:41
Subject Sample # 2	0.170	20:41
Breath Volume	2.132	20:41
Air Blank	0.000	20:42
Control Test	0.080	20:42
Air Blank	0.000	20:43
Diagnostics Check	OK	20:43

Data Download Date/Time: 12/19/2018 22:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREGORY,AMANDA,M
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: GREGORY,AMANDA
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 11/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: MCGUIRE

DATE/TIME OF TEST: 12/12/2018 21:03:28

Diagnostics Check	OK	20:56
Air Blank	0.000	20:57
Control Test	0.080	20:57
Air Blank	0.000	20:58
Subject Sample # 1	0.228	20:58
Breath Volume	2.363	20:58
Air Blank	0.000	20:59
Air Blank	0.000	21:01
Subject Sample # 2	0.213	21:01
Breath Volume	2.218	21:01
Air Blank	0.000	21:02
Control Test	0.080	21:02
Air Blank	0.000	21:03
Diagnostics Check	OK	21:03

Data Download Date/Time: 12/19/2018 22:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BREVIK,LONNIE,J
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: BREVIK,LONNIE
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 11/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: KOERSELMAN

DATE/TIME OF TEST: 12/13/2018 19:42:07

Diagnostics Check	OK	19:35
Air Blank	0.000	19:36
Control Test	0.080	19:36
Air Blank	0.000	19:37
Subject Sample # 1	0.186	19:37
Breath Volume	3.304	19:37
Air Blank	0.000	19:38
Air Blank	0.000	19:40
Subject Sample # 2	0.183	19:40
Breath Volume	2.996	19:40
Air Blank	0.000	19:41
Control Test	0.078	19:41
Air Blank	0.000	19:42
Diagnostics Check	OK	19:42

Data Download Date/Time: 12/19/2018 22:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUTY,J,K
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: CUTY,J
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 11/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:16

SUBJECT'S LAST NAME: CALIFANO

DATE/TIME OF TEST: 12/14/2018 21:44:40

Diagnostics Check	OK	21:38
Air Blank	0.000	21:39
Control Test	0.080	21:39
Air Blank	0.000	21:39
Subject Sample # 1	0.000	21:40
Breath Volume	2.363	21:40
Air Blank	0.000	21:40
Air Blank	0.000	21:42
Subject Sample # 2	0.000	21:43
Breath Volume	2.273	21:43
Air Blank	0.000	21:43
Control Test	0.078	21:44
Air Blank	0.000	21:44
Diagnostics Check	OK	21:44

Data Download Date/Time: 12/19/2018 22:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PECCI,Z,A
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: PECCI,Z
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 11/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:59

SUBJECT'S LAST NAME: SANDERS

DATE/TIME OF TEST: 12/14/2018 22:26:31

Diagnostics Check	OK	22:19
Air Blank	0.000	22:20
Control Test	0.079	22:20
Air Blank	0.000	22:21
Subject Sample # 1	0.228	22:21
Breath Volume	1.148	22:21
Air Blank	0.000	22:22
Air Blank	0.000	22:24
Subject Sample # 2	0.224	22:24
Breath Volume	1.625	22:24
Air Blank	0.000	22:25
Control Test	0.078	22:25
Air Blank	0.000	22:26
Diagnostics Check	OK	22:26

Data Download Date/Time: 12/19/2018 22:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGERVIL,CLAUDY,P
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: ANGERVIL,CLAUDY
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 11/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: MINOR

DATE/TIME OF TEST: 12/15/2018 00:12:27

Diagnostics Check	OK	00:05
Air Blank	0.000	00:06
Control Test	0.079	00:06
Air Blank	0.000	00:06
Subject Sample # 1	0.116	00:07
Breath Volume	1.699	00:07
Air Blank	0.000	00:08
Air Blank	0.000	00:10
Subject Sample # 2	0.114	00:10
Breath Volume	3.093	00:10
Air Blank	0.000	00:11
Control Test	0.078	00:11
Air Blank	0.000	00:12
Diagnostics Check	OK	00:12

Data Download Date/Time: 12/19/2018 22:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 11/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: CIFUENTES MENDOZA

DATE/TIME OF TEST: 12/15/2018 00:33:09

Diagnostics Check	OK	00:25
Air Blank	0.000	00:25
Control Test	0.079	00:25
Air Blank	0.000	00:26
Subject Sample # 1	0.242	00:27
Breath Volume	1.230	00:27
Air Blank	0.000	00:28
Air Blank	0.000	00:30
Subject Sample # 2	0.252	00:31
Breath Volume	2.871	00:31
Air Blank	0.000	00:32
Control Test	0.077	00:32
Air Blank	0.000	00:33
Diagnostics Check	OK	00:33

Data Download Date/Time: 12/19/2018 22:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLASZYK,JACOB,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BLASZYK,JACOB
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 11/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 12/15/2018 19:48:15

Diagnostics Check	OK	19:41
Air Blank	0.000	19:41
Control Test	0.080	19:42
Air Blank	0.000	19:42
Subject Sample # 1	0.105	19:43
Breath Volume	2.554	19:43
Air Blank	0.000	19:43
Air Blank	0.000	19:45
Subject Sample # 2	0.104	19:46
Breath Volume	2.875	19:46
Air Blank	0.000	19:47
Control Test	0.079	19:47
Air Blank	0.000	19:48
Diagnostics Check	OK	19:48

Data Download Date/Time: 12/19/2018 22:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGERVIL,CLAUDY,P
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: ANGERVIL,CLAUDY
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 11/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 12/16/2018 01:45:18

Diagnostics Check	OK	01:40	
Air Blank	0.000	01:41	
Control Test	0.079	01:41	
Air Blank	0.000	01:42	
Subject Sample	# 1	01:43	Subject Test Refused
Breath Volume	0.000	01:43	
Air Blank	0.000	01:44	
Control Test	0.079	01:44	
Air Blank	0.000	01:45	
Diagnostics Check	OK	01:45	

Data Download Date/Time: 12/19/2018 22:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLASZYK,JACOB,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BLASZYK,JACOB
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 11/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: HOLMES

DATE/TIME OF TEST: 12/16/2018 04:00:39

Diagnostics Check	OK	03:53
Air Blank	0.000	03:54
Control Test	0.080	03:54
Air Blank	0.000	03:55
Subject Sample # 1	0.154	03:55
Breath Volume	3.136	03:55
Air Blank	0.000	03:56
Air Blank	0.000	03:58
Subject Sample # 2	0.134	03:59
Breath Volume	1.246	03:59
Air Blank	0.000	03:59
Control Test	0.079	04:00
Air Blank	0.000	04:00
Diagnostics Check	OK	04:00

Data Download Date/Time: 12/19/2018 22:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 11/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: BAJANA-MOLINA

DATE/TIME OF TEST: 12/16/2018 21:22:23

Diagnostics Check	OK	21:10	
Air Blank	0.000	21:11	
Control Test	0.079	21:11	
Air Blank	0.000	21:12	
Subject Sample	# 1	21:15	No Sample Provided
Breath Volume	0.000	21:15	
Air Blank	0.000	21:15	
Air Blank	0.000	21:17	
Subject Sample	# 2	21:20	No Sample Provided
Breath Volume	0.000	21:20	
Air Blank	0.000	21:21	
Control Test	0.080	21:21	
Air Blank	0.000	21:22	
Diagnostics Check	OK	21:22	

Data Download Date/Time: 12/19/2018 22:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HATT,CODY,L
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: HATT,CODY
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 11/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: MCGONIGAL

DATE/TIME OF TEST: 12/17/2018 00:36:22

Diagnostics Check	OK	00:29
Air Blank	0.000	00:30
Control Test	0.079	00:30
Air Blank	0.000	00:31
Subject Sample # 1	0.158	00:31
Breath Volume	2.355	00:31
Air Blank	0.000	00:32
Air Blank	0.000	00:34
Subject Sample # 2	0.155	00:34
Breath Volume	3.484	00:34
Air Blank	0.000	00:35
Control Test	0.078	00:35
Air Blank	0.000	00:36
Diagnostics Check	OK	00:36

Data Download Date/Time: 12/19/2018 22:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAY,ERICK,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: DAY,ERICK
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 11/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 14:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:48

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 12/19/2018 15:36:07

Diagnostics Check	OK	15:29
Air Blank	0.000	15:30
Control Test	0.079	15:30
Air Blank	0.000	15:31
Subject Sample # 1	0.224	15:31
Breath Volume	1.910	15:31
Air Blank	0.000	15:32
Air Blank	0.000	15:34
Subject Sample # 2	0.230	15:34
Breath Volume	1.640	15:34
Air Blank	0.000	15:35
Control Test	0.079	15:35
Air Blank	0.000	15:36
Diagnostics Check	OK	15:36

Data Download Date/Time: 12/19/2018 22:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: LOGUE

DATE/TIME OF TEST: 12/22/2018 03:33:45

Diagnostics Check	OK	03:27
Air Blank	0.000	03:28
Control Test	0.081	03:28
Air Blank	0.000	03:28
Subject Sample # 1	0.172	03:29
Breath Volume	1.781	03:29
Air Blank	0.000	03:30
Air Blank	0.000	03:31
Subject Sample # 2	0.184	03:32
Breath Volume	1.988	03:32
Air Blank	0.000	03:32
Control Test	0.081	03:33
Air Blank	0.000	03:33
Diagnostics Check	OK	03:33

Data Download Date/Time: 01/13/2019 22:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUTY,J,K
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: CUTY,J
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:39

SUBJECT'S LAST NAME: BERGER

DATE/TIME OF TEST: 12/22/2018 21:15:17

Diagnostics Check	OK	20:59	
Air Blank	0.000	21:00	
Control Test	0.081	21:00	
Air Blank	0.000	21:01	
Subject Sample	# 1 0.187	21:02	
Breath Volume	1.773	21:02	
Air Blank	0.000	21:02	
Air Blank	0.000	21:04	
Subject Sample	# 2 0.185	21:07	Volume Not Met
Breath Volume	0.167	21:07	
Air Blank	0.000	21:08	
Air Blank	0.000	21:10	
Subject Sample	# 3 0.187	21:13	Volume Not Met
Breath Volume	0.234	21:13	
Air Blank	0.000	21:14	
Control Test	0.078	21:14	
Air Blank	0.000	21:15	
Diagnostics Check	OK	21:15	

Data Download Date/Time: 01/13/2019 22:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: VELASQUEZ-MENDEZ

DATE/TIME OF TEST: 12/25/2018 01:44:40

Diagnostics Check	OK	01:38
Air Blank	0.000	01:38
Control Test	0.080	01:38
Air Blank	0.000	01:39
Subject Sample # 1	0.259	01:39
Breath Volume	3.289	01:39
Air Blank	0.000	01:40
Air Blank	0.000	01:42
Subject Sample # 2	0.260	01:43
Breath Volume	3.554	01:43
Air Blank	0.000	01:43
Control Test	0.080	01:44
Air Blank	0.000	01:44
Diagnostics Check	OK	01:44

Data Download Date/Time: 01/13/2019 22:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLASZYK,JACOB,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BLASZYK,JACOB
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: WILEY

DATE/TIME OF TEST: 12/30/2018 22:58:12

Diagnostics Check	OK	22:51
Air Blank	0.000	22:52
Control Test	0.080	22:52
Air Blank	0.000	22:53
Subject Sample # 1	0.000	22:53
Breath Volume	3.031	22:53
Air Blank	0.000	22:54
Air Blank	0.000	22:56
Subject Sample # 2	0.000	22:56
Breath Volume	3.878	22:56
Air Blank	0.000	22:57
Control Test	0.080	22:57
Air Blank	0.000	22:58
Diagnostics Check	OK	22:58

Data Download Date/Time: 01/13/2019 22:58