Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVAREZ,BELINDA,R
BREATH TEST OPERATOR AGENCY: Aventura PD
ARRESTING OFFICER: GATTI,C
ARRESTING OFFICER AGENCY: Aventura PD
LAST AGENCY INSPECTION: 12/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: GOLDSMITH

DATE/TIME OF TEST: 01/01/2019 20:10:29

Diagnostics Check		OK	19:59	
Air Blank		0.000	19:59	
Control Test		0.078	19:59	
Air Blank		0.000	20:00	
Subject Sample	# 1		20:03	No Sample Provided
Breath Volume		0.000	20:03	
Air Blank		0.000	20:03	
Air Blank		0.000	20:05	
Subject Sample	# 2		20:09	No Sample Provided
Breath Volume		0.000	20:09	
Air Blank		0.000	20:09	
Control Test		0.077	20:09	
Air Blank		0.000	20:10	
Diagnostics Check		OK	20:10	

Data Download Date/Time: 01/11/2019 11:17

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KONETZ,JONATHAN,D
BREATH TEST OPERATOR AGENCY: Sunny Isles Beach PD
ARRESTING OFFICER: KONETZ,JONATHAN
ARRESTING OFFICER AGENCY: Sunny Isles Beach PD
LAST AGENCY INSPECTION: 12/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: DABNEY

DATE/TIME OF TEST: 01/07/2019 04:27:47

Diagnostics Check		OK	04:16	
Air Blank		0.000	04:16	
Control Test		0.077	04:17	
Air Blank		0.000	04:17	
Subject Sample	# 1		04:20	No Sample Provided
Breath Volume		0.000	04:20	
Air Blank		0.000	04:21	
Air Blank		0.000	04:23	
Subject Sample	# 2		04:26	No Sample Provided
Breath Volume		0.000	04:26	
Air Blank		0.000	04:26	
Control Test		0.078	04:27	
Air Blank		0.000	04:27	
Diagnostics Check		OK	04:27	

Data Download Date/Time: 01/11/2019 11:17

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KONETZ,JONATHAN,D
BREATH TEST OPERATOR AGENCY: Sunny Isles Beach PD
ARRESTING OFFICER: KONETZ,JONATHAN
ARRESTING OFFICER AGENCY: Sunny Isles Beach PD
LAST AGENCY INSPECTION: 12/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: DABNEY

DATE/TIME OF TEST: 01/07/2019 04:41:16

Diagnostics Check		OK	04:29	
Air Blank		0.000	04:30	
Control Test		0.077	04:30	
Air Blank		0.000	04:30	
Subject Sample	# 1		04:34	No Sample Provided
Breath Volume		0.000	04:34	
Air Blank		0.000	04:34	
Air Blank		0.000	04:36	
Subject Sample	# 2	0.144	04:39	Volume Not Met
Breath Volume		0.152	04:39	
Air Blank		0.000	04:40	
Control Test		0.079	04:40	
Air Blank		0.000	04:41	
Diagnostics Check		OK	04:41	

Data Download Date/Time: 01/11/2019 11:17

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KONETZ,JONATHAN,D
BREATH TEST OPERATOR AGENCY: Sunny Isles Beach PD
ARRESTING OFFICER: KONETZ,JONATHAN
ARRESTING OFFICER AGENCY: Sunny Isles Beach PD
LAST AGENCY INSPECTION: 12/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: DABNEY

DATE/TIME OF TEST: 01/07/2019 04:59:20

Diagnostics Check		OK	04:43	
Air Blank		0.000	04:43	
Control Test		0.076	04:43	
Air Blank		0.000	04:44	
Subject Sample	# 1	0.134	04:46	
Breath Volume		1.324	04:46	
Air Blank		0.000	04:47	
Air Blank		0.000	04:49	
Subject Sample	# 2		04:52	No Sample Provided
Breath Volume		0.000	04:52	
Air Blank		0.000	04:52	
Air Blank		0.000	04:54	
Subject Sample	# 3		04:57	No Sample Provided
Breath Volume		0.000	04:57	
Air Blank		0.000	04:58	
Control Test		0.077	04:58	
Air Blank		0.000	04:59	
Diagnostics Check		OK	04:59	

Data Download Date/Time: 01/11/2019 11:17

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORGES,SEAN,P
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: BORGES,SEAN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 12/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 14:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:17

SUBJECT'S LAST NAME: TIRONE MOREAN

DATE/TIME OF TEST: 01/01/2019 14:41:40

Diagnostics Check		OK	14:38	
Air Blank		0.000	14:38	
Control Test		0.080	14:39	
Air Blank		0.000	14:39	
Subject Sample	# 1		14:40	Subject Test Refused
Breath Volume		0.000	14:40	
Air Blank		0.000	14:40	
Control Test		0.080	14:41	
Air Blank		0.000	14:41	
Diagnostics Check		OK	14:41	

Data Download Date/Time: 01/04/2019 10:31

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORGES,SEAN,P
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: BORGES,SEAN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 07:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:22

SUBJECT'S LAST NAME: NAVARRO

DATE/TIME OF TEST: 01/09/2019 08:47:44

Diagnostics Check		OK	08:44	
Air Blank		0.000	08:45	
Control Test		0.080	08:45	
Air Blank		0.000	08:46	
Subject Sample	# 1		08:46	Subject Test Refused
Breath Volume		0.000	08:46	
Air Blank		0.000	08:46	
Control Test		0.080	08:47	
Air Blank		0.000	08:47	
Diagnostics Check		OK	08:47	

Data Download Date/Time: 02/08/2019 11:03

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUAREZ,CARLOS,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: BEHNEY,DAVID
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: MAYORGA-GARAY

DATE/TIME OF TEST: 01/10/2019 00:19:30

Diagnostics Check		OK	00:14	
Air Blank		0.000	00:15	
Control Test		0.080	00:15	
Air Blank		0.000	00:16	
Subject Sample	# 1		00:18	Subject Test Refused
Breath Volume		0.000	00:18	
Air Blank		0.000	00:18	
Control Test		0.080	00:18	
Air Blank		0.000	00:19	
Diagnostics Check		OK	00:19	

Data Download Date/Time: 02/08/2019 11:04

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORENO,FRANCISCO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: MORENO,FRANCISCO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 08:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:33

SUBJECT'S LAST NAME: SIERRA

DATE/TIME OF TEST: 01/10/2019 10:00:37

Diagnostics Check		OK	09:53
Air Blank		0.000	09:54
Control Test		0.080	09:54
Air Blank		0.000	09:55
Subject Sample	# 1	0.201	09:55
Breath Volume		1.226	09:55
Air Blank		0.000	09:56
Air Blank		0.000	09:58
Subject Sample	# 2	0.206	09:58
Breath Volume		1.203	09:58
Air Blank		0.000	09:59
Control Test		0.078	10:00
Air Blank		0.000	10:00
Diagnostics Check		OK	10:00

Data Download Date/Time: 02/08/2019 11:04

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTILLO,JOSUET,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CASTILLO,JOSUET
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: SONCK

DATE/TIME OF TEST: 01/11/2019 01:56:29

Diagnostics Check		OK	01:48
Air Blank		0.000	01:48
Control Test		0.081	01:48
Air Blank		0.000	01:49
Subject Sample	# 1	0.081	01:51
Breath Volume		1.226	01:51
Air Blank		0.000	01:52
Air Blank		0.000	01:53
Subject Sample	# 2	0.080	01:54
Breath Volume		1.296	01:54
Air Blank		0.000	01:55
Control Test		0.078	01:55
Air Blank		0.000	01:56
Diagnostics Check		OK	01:56

Data Download Date/Time: 02/08/2019 11:04

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLORIAN,ROLANDO,I
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FLORIAN,ROLANDO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: GONZALEZ BENITEZ

DATE/TIME OF TEST: 01/11/2019 21:14:21

Diagnostics Check		OK	21:11	
Air Blank		0.000	21:11	
Control Test		0.080	21:12	
Air Blank		0.000	21:12	
Subject Sample	# 1		21:12	Subject Test Refused
Breath Volume		0.000	21:12	
Air Blank		0.000	21:13	
Control Test		0.080	21:13	
Air Blank		0.000	21:14	
Diagnostics Check		OK	21:14	

Data Download Date/Time: 02/08/2019 11:04

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLORIAN,ROLANDO,I
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FLORIAN,ROLANDO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: GONZALEZ BENITEZ

DATE/TIME OF TEST: 01/11/2019 21:54:19

Diagnostics Check		OK	21:48	
Air Blank		0.000	21:49	
Control Test		0.081	21:49	
Air Blank		0.000	21:50	
Subject Sample	# 1	0.139	21:52	
Breath Volume		1.402	21:52	
Air Blank			21:53	Purge Fail
Air Blank		0.000	21:54	

Data Download Date/Time: 02/08/2019 11:04

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLORIAN,ROLANDO,I
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FLORIAN,ROLANDO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: GONZALEZ BENITEZ

DATE/TIME OF TEST: 01/11/2019 22:01:05

Diagnostics Check		OK	21:56	
Air Blank		0.000	21:57	
Control Test		0.076	21:57	
Air Blank		0.000	21:57	
Subject Sample	# 1	0.130	21:59	
Breath Volume		1.277	21:59	
Air Blank			22:00	Purge Fail
Air Blank		0.000	22:01	

Data Download Date/Time: 02/08/2019 11:04

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLORIAN,ROLANDO,I
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FLORIAN,ROLANDO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: GONZALEZ BENITEZ

DATE/TIME OF TEST: 01/11/2019 22:05:36

Diagnostics Check		OK	22:02	
Air Blank		0.000	22:02	
Control Test		0.075	22:03	
Air Blank		0.000	22:03	
Subject Sample	# 1	0.144	22:04	
Breath Volume		3.406	22:04	
Air Blank			22:05	Purge Fail
Air Blank		0.000	22:05	

Data Download Date/Time: 02/08/2019 11:04

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLORIAN,ROLANDO,I
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FLORIAN,ROLANDO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: GONZALEZ BENITEZ

DATE/TIME OF TEST: 01/11/2019 22:08:20

Diagnostics Check	OK	22:06	
Air Blank	0.000	22:07	
Control Test	0.074	22:07	Control Outside Tolerance
Air Blank	0.000	22:08	

Data Download Date/Time: 02/08/2019 11:04

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLORIAN,ROLANDO,I
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FLORIAN,ROLANDO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: GONZALEZ BENITEZ

DATE/TIME OF TEST: 01/11/2019 22:17:44

Diagnostics Check		OK	22:09
Air Blank		0.000	22:10
Control Test		0.078	22:10
Air Blank		0.000	22:11
Subject Sample	# 1	0.129	22:12
Breath Volume		1.800	22:12
Air Blank		0.000	22:13
Air Blank		0.000	22:15
Subject Sample	# 2	0.126	22:15
Breath Volume		1.953	22:15
Air Blank		0.000	22:16
Control Test		0.075	22:17
Air Blank		0.000	22:17
Diagnostics Check		OK	22:17

Data Download Date/Time: 02/08/2019 11:04

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUJILLO,GEORGE,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: TRUJILLO,GEORGE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: ALBURY

DATE/TIME OF TEST: 01/11/2019 22:28:09

Diagnostics Check		OK	22:22	
Air Blank		0.000	22:23	
Control Test		0.077	22:23	
Air Blank		0.000	22:24	
Subject Sample	# 1	0.231	22:26	
Breath Volume		1.566	22:26	
Air Blank			22:27	Purge Fail
Air Blank		0.000	22:28	

Data Download Date/Time: 02/08/2019 11:04

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUJILLO,GEORGE,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: TRUJILLO,GEORGE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: ALBURY

DATE/TIME OF TEST: 01/11/2019 22:31:20

Diagnostics Check	OK	22:29	
Air Blank	0.000	22:30	
Control Test	0.073	22:30	Control Outside Tolerance
Air Blank	0.000	22:31	

Data Download Date/Time: 02/08/2019 11:04

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUJILLO,GEORGE,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: TRUJILLO,GEORGE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: ALBURY

DATE/TIME OF TEST: 01/11/2019 22:42:25

Diagnostics Check		OK	22:33
Air Blank		0.000	22:33
Control Test		0.079	22:33
Air Blank		0.000	22:34
Subject Sample	# 1	0.241	22:34
Breath Volume		1.269	22:34
Air Blank		0.000	22:35
Air Blank		0.000	22:37
Subject Sample	# 2	0.220	22:37
Breath Volume		1.242	22:37
Air Blank		0.000	22:38
Air Blank		0.000	22:40
Subject Sample	# 3	0.237	22:40
Breath Volume		1.437	22:40
Air Blank		0.000	22:41
Control Test		0.076	22:41
Air Blank		0.000	22:42
Diagnostics Check		OK	22:42

Data Download Date/Time: 02/08/2019 11:04

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIDDO,RAAJU,S
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: SIDDO,RAAJU
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: LOPEZ PERALTA

DATE/TIME OF TEST: 01/12/2019 00:17:06

Diagnostics Check		OK	00:11	
Air Blank		0.000	00:11	
Control Test		0.078	00:12	
Air Blank		0.000	00:12	
Subject Sample	# 1	0.157	00:15	Volume Not Met
Breath Volume		0.628	00:15	
Air Blank			00:16	Purge Fail
Air Blank		0.000	00:17	

Data Download Date/Time: 02/08/2019 11:04

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIDDO,RAAJU,S
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: SIDDO,RAAJUS
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: LOPEZ PERALTA

DATE/TIME OF TEST: 01/12/2019 00:30:47

Diagnostics Check		OK	00:18	
Air Blank		0.000	00:19	
Control Test		0.076	00:19	
Air Blank		0.000	00:20	
Subject Sample	# 1	0.127	00:23	Volume Not Met
Breath Volume		0.113	00:23	
Air Blank		0.000	00:24	
Air Blank		0.000	00:25	
Subject Sample	# 2	0.165	00:28	Slope Not Level
Breath Volume		1.492	00:28	
Air Blank		0.000	00:29	
Control Test		0.066	00:30	Control Outside Tolerance
Air Blank		0.000	00:30	

Data Download Date/Time: 02/08/2019 11:04

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIDDO,RAAJU,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: SIDDO,RAAJU
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: LOPEZ PERALTA

DATE/TIME OF TEST: 01/12/2019 00:37:40

Diagnostics Check		OK	00:33	
Air Blank		0.000	00:33	
Control Test		0.076	00:33	
Air Blank		0.000	00:34	
Subject Sample	# 1		00:35	Subject Test Refused
Breath Volume		0.171	00:35	
Air Blank		0.000	00:36	
Control Test		0.078	00:37	
Air Blank		0.000	00:37	
Diagnostics Check		OK	00:37	

Data Download Date/Time: 02/08/2019 11:04

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,D,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VARGAS,D
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 01/12/2019 03:39:16

Diagnostics Check		OK	03:32
Air Blank		0.000	03:32
Control Test		0.078	03:33
Air Blank		0.000	03:33
Subject Sample	# 1	0.136	03:34
Breath Volume		2.734	03:34
Air Blank		0.000	03:35
Air Blank		0.000	03:37
Subject Sample	# 2	0.131	03:37
Breath Volume		1.789	03:37
Air Blank		0.000	03:38
Control Test		0.079	03:38
Air Blank		0.000	03:39
Diagnostics Check		OK	03:39

Data Download Date/Time: 02/08/2019 11:04

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,D,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VARGAS,D
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: FRIMAN CONTRERAS

DATE/TIME OF TEST: 01/13/2019 04:18:59

Diagnostics Check		OK	04:11
Air Blank		0.000	04:12
Control Test		0.079	04:12
Air Blank		0.000	04:13
Subject Sample	# 1	0.210	04:14
Breath Volume		1.195	04:14
Air Blank		0.000	04:15
Air Blank		0.000	04:16
Subject Sample	# 2	0.230	04:17
Breath Volume		1.449	04:17
Air Blank		0.000	04:18
Control Test		0.078	04:18
Air Blank		0.000	04:18
Diagnostics Check		OK	04:18

Data Download Date/Time: 02/08/2019 11:04

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTILLO,JOSUET,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CASTILLO,JOSUET
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: MONTERO

DATE/TIME OF TEST: 01/13/2019 04:27:36

Diagnostics Check		OK	04:22	
Air Blank		0.000	04:23	
Control Test		0.078	04:23	
Air Blank		0.000	04:23	
Subject Sample	# 1		04:26	Subject Test Refused
Breath Volume		0.000	04:26	
Air Blank		0.000	04:26	
Control Test		0.079	04:26	
Air Blank		0.000	04:27	
Diagnostics Check		OK	04:27	

Data Download Date/Time: 02/08/2019 11:04

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,D,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: DELGADO,M
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 01/13/2019 05:06:04

Diagnostics Check		OK	05:03	
Air Blank		0.000	05:03	
Control Test		0.079	05:03	
Air Blank		0.000	05:04	
Subject Sample	# 1		05:04	Subject Test Refused
Breath Volume		0.000	05:04	
Air Blank		0.000	05:05	
Control Test		0.080	05:05	
Air Blank		0.000	05:05	
Diagnostics Check		OK	05:06	

Data Download Date/Time: 02/08/2019 11:04

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,D,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: RAWLUK,S
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:58

SUBJECT'S LAST NAME: MORENO RODRIGUEZ

DATE/TIME OF TEST: 01/13/2019 06:08:18

Diagnostics Check		OK	06:05	
Air Blank		0.000	06:05	
Control Test		0.080	06:06	
Air Blank		0.000	06:06	
Subject Sample	# 1		06:06	Subject Test Refused
Breath Volume		0.000	06:06	
Air Blank		0.000	06:07	
Control Test		0.080	06:07	
Air Blank		0.000	06:08	
Diagnostics Check		OK	06:08	

Data Download Date/Time: 02/08/2019 11:04

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUERTA,EDWIN,
BREATH TEST OPERATOR AGENCY: Hialeah Gardens PD
ARRESTING OFFICER: HUERTA,EDWIN
ARRESTING OFFICER AGENCY: Hialeah Gardens PD
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 05:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:35

SUBJECT'S LAST NAME: PORTELA CRESPO

DATE/TIME OF TEST: 01/13/2019 06:34:26

Diagnostics Check		OK	06:28
Air Blank		0.000	06:28
Control Test		0.079	06:28
Air Blank		0.000	06:29
Subject Sample	# 1	0.250	06:29
Breath Volume		2.921	06:29
Air Blank		0.000	06:30
Air Blank		0.000	06:32
Subject Sample	# 2	0.258	06:32
Breath Volume		2.734	06:32
Air Blank		0.000	06:33
Control Test		0.079	06:33
Air Blank		0.000	06:34
Diagnostics Check		OK	06:34

Data Download Date/Time: 02/08/2019 11:04

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NADAL,JOSE,V
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: NADAL,JOSE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: CORREA LOPEZ

DATE/TIME OF TEST: 01/16/2019 02:26:38

Diagnostics Check		OK	02:19
Air Blank		0.000	02:20
Control Test		0.081	02:20
Air Blank		0.000	02:21
Subject Sample	# 1	0.298	02:21
Breath Volume		2.230	02:21
Air Blank		0.000	02:22
Air Blank		0.000	02:24
Subject Sample	# 2	0.305	02:24
Breath Volume		2.132	02:24
Air Blank		0.000	02:25
Control Test		0.078	02:26
Air Blank		0.000	02:26
Diagnostics Check		OK	02:26

Data Download Date/Time: 02/08/2019 11:04

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVALOS,DANIEL,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: GONZALEZ,JOSE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: PADRON DUENAS

DATE/TIME OF TEST: 01/16/2019 02:57:26

Diagnostics Check		OK	02:53	
Air Blank		0.000	02:54	
Control Test		0.079	02:54	
Air Blank		0.000	02:55	
Subject Sample	# 1		02:56	Subject Test Refused
Breath Volume		0.000	02:56	
Air Blank		0.000	02:56	
Control Test		0.080	02:56	
Air Blank		0.000	02:57	
Diagnostics Check		OK	02:57	

Data Download Date/Time: 02/08/2019 11:04

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: GUERRA

DATE/TIME OF TEST: 01/19/2019 03:49:56

Diagnostics Check		OK	03:41
Air Blank		0.000	03:42
Control Test		0.080	03:42
Air Blank		0.000	03:43
Subject Sample	# 1	0.085	03:43
Breath Volume		1.402	03:43
Air Blank		0.000	03:44
Air Blank		0.000	03:46
Subject Sample	# 2	0.094	03:48
Breath Volume		1.886	03:48
Air Blank		0.000	03:49
Control Test		0.080	03:49
Air Blank		0.000	03:49
Diagnostics Check		OK	03:49

Data Download Date/Time: 02/08/2019 11:04

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,D,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VARGAS,D
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 01/19/2019 04:40:52

Diagnostics Check		OK	04:27	
Air Blank		0.000	04:28	
Control Test		0.079	04:28	
Air Blank		0.000	04:29	
Subject Sample	# 1	0.158	04:30	
Breath Volume		1.308	04:30	
Air Blank		0.000	04:31	
Air Blank		0.000	04:32	
Subject Sample	# 2	0.119	04:35	Volume Not Met
Breath Volume		0.648	04:35	
Air Blank		0.000	04:36	
Air Blank		0.000	04:38	
Subject Sample	# 3	0.153	04:39	
Breath Volume		1.457	04:39	
Air Blank		0.000	04:39	
Control Test		0.077	04:40	
Air Blank		0.000	04:40	
Diagnostics Check		OK	04:40	

Data Download Date/Time: 02/08/2019 11:04

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUJILLO,GEORGE,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: TRUJILLO,GEORGE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: NO DL

DATE/TIME OF TEST: 01/19/2019 20:00:00

Diagnostics Check		OK	19:57	
Air Blank		0.000	19:57	
Control Test		0.080	19:57	
Air Blank		0.000	19:58	
Subject Sample	# 1		19:58	Subject Test Refused
Breath Volume		0.000	19:58	
Air Blank		0.000	19:59	
Control Test		0.080	19:59	
Air Blank		0.000	19:59	
Diagnostics Check		OK	20:00	

Data Download Date/Time: 02/08/2019 11:04

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCAULEY,DEVERO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: SOTO CABALLERO,ALEXEI
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: ORTA

DATE/TIME OF TEST: 01/19/2019 22:23:22

Diagnostics Check		OK	22:11	
Air Blank		0.000	22:11	
Control Test		0.081	22:11	
Air Blank		0.000	22:12	
Subject Sample	# 1		22:15	No Sample Provided
Breath Volume		0.000	22:15	
Air Blank		0.000	22:15	
Air Blank		0.000	22:17	
Subject Sample	# 2	0.217	22:18	No .020 agreement
Breath Volume		2.476	22:18	
Air Blank		0.000	22:19	
Air Blank		0.000	22:21	
Subject Sample	# 3	0.192	22:21	No .020 agreement
Breath Volume		1.707	22:21	
Air Blank		0.000	22:22	
Control Test		0.078	22:22	
Air Blank		0.000	22:23	
Diagnostics Check		OK	22:23	

Data Download Date/Time: 02/08/2019 11:04

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCAULEY,DEVERO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: SOTO CABALLERO,ALEXEI
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: ORTA

DATE/TIME OF TEST: 01/19/2019 22:37:48

Diagnostics Check		OK	22:27
Air Blank		0.000	22:28
Control Test		0.079	22:28
Air Blank		0.000	22:29
Subject Sample	# 1	0.210	22:29
Breath Volume		2.035	22:29
Air Blank		0.000	22:30
Air Blank		0.000	22:32
Subject Sample	# 2	0.189	22:32
Breath Volume		1.707	22:32
Air Blank		0.000	22:33
Air Blank		0.000	22:35
Subject Sample	# 3	0.199	22:35
Breath Volume		3.406	22:35
Air Blank		0.000	22:36
Control Test		0.077	22:37
Air Blank		0.000	22:37
Diagnostics Check		OK	22:37

Data Download Date/Time: 02/08/2019 11:04

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUAN,RALPH,B
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: CUAN,RALPH
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:23

SUBJECT'S LAST NAME: RESTREPO

DATE/TIME OF TEST: 01/20/2019 17:56:16

Diagnostics Check		OK	17:53	
Air Blank		0.000	17:53	
Control Test		0.079	17:54	
Air Blank		0.000	17:54	
Subject Sample	# 1		17:54	Subject Test Refused
Breath Volume		0.000	17:54	
Air Blank		0.000	17:55	
Control Test		0.079	17:55	
Air Blank		0.000	17:56	
Diagnostics Check		OK	17:56	

Data Download Date/Time: 02/08/2019 11:04

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLORIAN,ROLANDO,I
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FLORIAN,ROLANDO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: GONZALEZ RODRIGUEZ

DATE/TIME OF TEST: 01/21/2019 21:30:31

Diagnostics Check		OK	21:16	
Air Blank		0.000	21:16	
Control Test		0.080	21:17	
Air Blank		0.000	21:17	
Subject Sample	# 1		21:20	No Sample Provided
Breath Volume		0.000	21:20	
Air Blank		0.000	21:21	
Air Blank		0.000	21:23	
Subject Sample	# 2	0.212	21:25	
Breath Volume		2.238	21:25	
Air Blank		0.000	21:26	
Air Blank		0.000	21:27	
Subject Sample	# 3		21:28	Slope Not Met
Breath Volume		1.066	21:28	
Air Blank		0.000	21:29	
Control Test		0.076	21:29	
Air Blank		0.000	21:30	
Diagnostics Check		OK	21:30	

Data Download Date/Time: 02/08/2019 11:04

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLORIAN,ROLANDO,I
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FLORIAN,ROLANDO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: GONZALEZ RODRIGUEZ

DATE/TIME OF TEST: 01/21/2019 21:38:56

Diagnostics Check		OK	21:32	
Air Blank		0.000	21:32	
Control Test		0.077	21:32	
Air Blank		0.000	21:33	
Subject Sample	# 1	0.232	21:33	
Breath Volume		1.367	21:33	
Air Blank		0.000	21:34	
Air Blank		0.000	21:36	
Subject Sample	# 2	0.216	21:37	
Breath Volume		1.867	21:37	
Air Blank			21:38	Purge Fail
Air Blank		0.000	21:38	

Data Download Date/Time: 02/08/2019 11:04

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLORIAN,ROLANDO,I
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FLORIAN,ROLANDO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: GONZALEZ RODRIGUEZ

DATE/TIME OF TEST: 01/21/2019 21:41:34

Diagnostics Check	OK	21:40	
Air Blank	0.000	21:40	
Control Test	0.073	21:40	Control Outside Tolerance
Air Blank	0.000	21:41	

Data Download Date/Time: 02/08/2019 11:04

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLORIAN,ROLANDO,I
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FLORIAN,ROLANDO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: GONZALEZ RODRIGUEZ

DATE/TIME OF TEST: 01/21/2019 21:46:58

Diagnostics Check		OK	21:43	
Air Blank		0.000	21:44	
Control Test		0.076	21:44	
Air Blank		0.000	21:44	
Subject Sample	# 1	0.232	21:45	
Breath Volume		2.578	21:45	
Air Blank			21:46	Purge Fail
Air Blank		0.000	21:46	

Data Download Date/Time: 02/08/2019 11:04

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLORIAN,ROLANDO,I
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FLORIAN,ROLANDO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: GONZALEZ RODRIGUEZ

DATE/TIME OF TEST: 01/21/2019 21:49:33

Diagnostics Check	OK	21:48	
Air Blank	0.000	21:48	
Control Test	0.074	21:48	Control Outside Tolerance
Air Blank	0.000	21:49	

Data Download Date/Time: 02/08/2019 11:04

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLORIAN,ROLANDO,I
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FLORIAN,ROLANDO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: GONZALEZ RODRIGUEZ

DATE/TIME OF TEST: 01/21/2019 21:58:40

Diagnostics Check		OK	21:51
Air Blank		0.000	21:52
Control Test		0.078	21:52
Air Blank		0.000	21:52
Subject Sample	# 1	0.229	21:53
Breath Volume		2.335	21:53
Air Blank		0.000	21:54
Air Blank		0.000	21:56
Subject Sample	# 2	0.236	21:56
Breath Volume		2.582	21:56
Air Blank		0.000	21:57
Control Test		0.076	21:58
Air Blank		0.000	21:58
Diagnostics Check		OK	21:58

Data Download Date/Time: 02/08/2019 11:04

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLORIAN,ROLANDO,I
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ORTEGA,IDALMYS
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: LAPRADD

DATE/TIME OF TEST: 01/21/2019 22:18:05

Diagnostics Check		OK	22:10
Air Blank		0.000	22:11
Control Test		0.078	22:11
Air Blank		0.000	22:12
Subject Sample	# 1	0.190	22:13
Breath Volume		3.191	22:13
Air Blank		0.000	22:14
Air Blank		0.000	22:15
Subject Sample	# 2	0.185	22:16
Breath Volume		1.183	22:16
Air Blank		0.000	22:17
Control Test		0.077	22:17
Air Blank		0.000	22:18
Diagnostics Check		OK	22:18

Data Download Date/Time: 02/08/2019 11:04

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: VALINA ALMENTERO

DATE/TIME OF TEST: 01/23/2019 01:53:18

Diagnostics Check		OK	01:50	
Air Blank		0.000	01:50	
Control Test		0.080	01:50	
Air Blank		0.000	01:51	
Subject Sample	# 1		01:51	Subject Test Refused
Breath Volume		0.000	01:51	
Air Blank		0.000	01:52	
Control Test		0.080	01:52	
Air Blank		0.000	01:53	
Diagnostics Check		OK	01:53	

Data Download Date/Time: 02/08/2019 11:04

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORENZO,YOSNIEL,
BREATH TEST OPERATOR AGENCY: Sweetwater PD
ARRESTING OFFICER: LORENZO,YOSNIEL
ARRESTING OFFICER AGENCY: Sweetwater PD
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: PEREZ ALONSO

DATE/TIME OF TEST: 01/23/2019 22:53:33

Diagnostics Check		OK	22:46
Air Blank		0.000	22:47
Control Test		0.080	22:47
Air Blank		0.000	22:48
Subject Sample	# 1	0.228	22:48
Breath Volume		1.375	22:48
Air Blank		0.000	22:49
Air Blank		0.000	22:51
Subject Sample	# 2	0.227	22:51
Breath Volume		2.113	22:51
Air Blank		0.000	22:52
Control Test		0.078	22:52
Air Blank		0.000	22:53
Diagnostics Check		OK	22:53

Data Download Date/Time: 02/08/2019 11:04

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: BRUNER

DATE/TIME OF TEST: 01/24/2019 01:39:18

Diagnostics Check		OK	01:33
Air Blank		0.000	01:33
Control Test		0.080	01:34
Air Blank		0.000	01:34
Subject Sample	# 1	0.000	01:35
Breath Volume		1.984	01:35
Air Blank		0.000	01:35
Air Blank		0.000	01:37
Subject Sample	# 2	0.000	01:37
Breath Volume		2.308	01:37
Air Blank		0.000	01:38
Control Test		0.080	01:38
Air Blank		0.000	01:39
Diagnostics Check		OK	01:39

Data Download Date/Time: 02/08/2019 11:04

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STEVEN,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: RIVERO,NICHOLAS
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: KIMBALL

DATE/TIME OF TEST: 01/24/2019 02:31:18

Diagnostics Check		OK	02:20
Air Blank		0.000	02:21
Control Test		0.080	02:21
Air Blank		0.000	02:22
Subject Sample	# 1	0.146	02:22
Breath Volume		1.214	02:22
Air Blank		0.000	02:23
Air Blank		0.000	02:25
Subject Sample	# 2	0.171	02:26
Breath Volume		2.214	02:26
Air Blank		0.000	02:26
Air Blank		0.000	02:28
Subject Sample	# 3	0.140	02:29
Breath Volume		1.160	02:29
Air Blank		0.000	02:30
Control Test		0.079	02:30
Air Blank		0.000	02:31
Diagnostics Check		OK	02:31

Data Download Date/Time: 02/08/2019 11:05

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PACHECO,GEORGE,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PACHECO,GEORGE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:13

SUBJECT'S LAST NAME: ABRANTE

DATE/TIME OF TEST: 01/24/2019 05:45:49

Diagnostics Check		OK	05:42	
Air Blank		0.000	05:42	
Control Test		0.080	05:42	
Air Blank		0.000	05:43	
Subject Sample	# 1		05:44	Subject Test Refused
Breath Volume		0.000	05:44	
Air Blank		0.000	05:44	
Control Test		0.080	05:45	
Air Blank		0.000	05:45	
Diagnostics Check		OK	05:45	

Data Download Date/Time: 02/08/2019 11:05

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PACHECO,GEORGE,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PACHECO,GEORGE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 05:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:50

SUBJECT'S LAST NAME: GOLINDANO DIAZ

DATE/TIME OF TEST: 01/25/2019 06:19:35

Diagnostics Check		OK	06:16	
Air Blank		0.000	06:17	
Control Test		0.080	06:17	
Air Blank		0.000	06:17	
Subject Sample	# 1		06:18	Subject Test Refused
Breath Volume		0.000	06:18	
Air Blank		0.000	06:18	
Control Test		0.080	06:18	
Air Blank		0.000	06:19	
Diagnostics Check		OK	06:19	

Data Download Date/Time: 02/08/2019 11:05

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARDELAS,MAYLEE,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: BARDELAS,MAYLE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: CERRUD

DATE/TIME OF TEST: 01/27/2019 05:28:36

Diagnostics Check		OK	05:25	
Air Blank		0.000	05:26	
Control Test		0.081	05:26	
Air Blank		0.000	05:26	
Subject Sample	# 1		05:27	Subject Test Refused
Breath Volume		0.000	05:27	
Air Blank		0.000	05:27	
Control Test		0.080	05:27	
Air Blank		0.000	05:28	
Diagnostics Check		OK	05:28	

Data Download Date/Time: 02/08/2019 11:05

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PACHECO,GEORGE,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PACHECO,GEORGE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 06:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:30

SUBJECT'S LAST NAME: GOMEZ OSORTO

DATE/TIME OF TEST: 01/27/2019 08:10:13

Diagnostics Check		OK	08:01
Air Blank		0.000	08:02
Control Test		0.081	08:02
Air Blank		0.000	08:02
Subject Sample	# 1	0.170	08:05
Breath Volume		2.128	08:05
Air Blank		0.000	08:06
Air Blank		0.000	08:07
Subject Sample	# 2	0.174	08:08
Breath Volume		1.937	08:08
Air Blank		0.000	08:09
Control Test		0.077	08:09
Air Blank		0.000	08:10
Diagnostics Check		OK	08:10

Data Download Date/Time: 02/08/2019 11:05

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NADAL,JOSE,V
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: NADAL,JOSE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:25

SUBJECT'S LAST NAME: SOMANO

DATE/TIME OF TEST: 01/29/2019 05:50:25

Diagnostics Check		OK	05:46	
Air Blank		0.000	05:46	
Control Test		0.081	05:46	
Air Blank		0.000	05:47	
Subject Sample	# 1		05:49	Subject Test Refused
Breath Volume		0.000	05:49	
Air Blank		0.000	05:49	
Control Test		0.080	05:49	
Air Blank		0.000	05:50	
Diagnostics Check		OK	05:50	

Data Download Date/Time: 02/08/2019 11:05

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PACHECO,GEORGE,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PACHECO,GEORGE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 01/30/2019 02:32:07

Diagnostics Check		OK	02:29	
Air Blank		0.000	02:29	
Control Test		0.080	02:30	
Air Blank		0.000	02:30	
Subject Sample	# 1		02:30	Subject Test Refused
Breath Volume		0.000	02:30	
Air Blank		0.000	02:31	
Control Test		0.081	02:31	
Air Blank		0.000	02:32	
Diagnostics Check		OK	02:32	

Data Download Date/Time: 02/08/2019 11:05

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTANA,JORGE,F
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VARGAS,D
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: SALAZAR

DATE/TIME OF TEST: 01/31/2019 03:45:18

Diagnostics Check		OK	03:38
Air Blank		0.000	03:38
Control Test		0.080	03:39
Air Blank		0.000	03:39
Subject Sample	# 1	0.237	03:40
Breath Volume		2.289	03:40
Air Blank		0.000	03:41
Air Blank		0.000	03:43
Subject Sample	# 2	0.227	03:43
Breath Volume		1.320	03:43
Air Blank		0.000	03:44
Control Test		0.076	03:44
Air Blank		0.000	03:45
Diagnostics Check		OK	03:45

Data Download Date/Time: 02/08/2019 11:05

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTANA,JORGE,F
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VARGAS,D
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: SALAZAR

DATE/TIME OF TEST: 01/31/2019 03:54:00

Diagnostics Check		OK	03:47
Air Blank		0.000	03:47
Control Test		0.079	03:48
Air Blank		0.000	03:48
Subject Sample	# 1	0.230	03:49
Breath Volume		1.687	03:49
Air Blank		0.000	03:50
Air Blank		0.000	03:51
Subject Sample	# 2	0.230	03:52
Breath Volume		1.332	03:52
Air Blank		0.000	03:53
Control Test		0.078	03:53
Air Blank		0.000	03:53
Diagnostics Check		OK	03:54

Data Download Date/Time: 02/08/2019 11:05

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:09

SUBJECT'S LAST NAME: LAFEVER

DATE/TIME OF TEST: 01/08/2019 19:37:27

Diagnostics Check		OK	19:31
Air Blank		0.000	19:31
Control Test		0.080	19:31
Air Blank		0.000	19:32
Subject Sample	# 1	0.114	19:32
Breath Volume		2.214	19:32
Air Blank		0.000	19:33
Air Blank		0.000	19:35
Subject Sample	# 2	0.112	19:35
Breath Volume		2.078	19:35
Air Blank		0.000	19:36
Control Test		0.080	19:36
Air Blank		0.000	19:37
Diagnostics Check		OK	19:37

Data Download Date/Time: 01/22/2019 13:47

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRYANT,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: HUGHES

DATE/TIME OF TEST: 01/09/2019 22:28:48

Diagnostics Check		OK	22:21
Air Blank		0.000	22:21
Control Test		0.081	22:21
Air Blank		0.000	22:22
Subject Sample	# 1	0.245	22:23
Breath Volume		1.277	22:23
Air Blank		0.000	22:23
Air Blank		0.000	22:25
Subject Sample	# 2	0.234	22:27
Breath Volume		1.785	22:27
Air Blank		0.000	22:27
Control Test		0.081	22:28
Air Blank		0.000	22:28
Diagnostics Check		OK	22:28

Data Download Date/Time: 01/22/2019 13:47

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,D,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TURPIN,D
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: WANLISS

DATE/TIME OF TEST: 01/11/2019 04:06:35

Diagnostics Check		OK	04:00
Air Blank		0.000	04:00
Control Test		0.081	04:00
Air Blank		0.000	04:01
Subject Sample	# 1	0.236	04:01
Breath Volume		3.539	04:01
Air Blank		0.000	04:02
Air Blank		0.000	04:04
Subject Sample	# 2	0.230	04:05
Breath Volume		4.261	04:05
Air Blank		0.000	04:05
Control Test		0.081	04:06
Air Blank		0.000	04:06
Diagnostics Check		OK	04:06

Data Download Date/Time: 01/22/2019 13:47

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAGE,D,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CENSON,T
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:06

SUBJECT'S LAST NAME: DOWNING

DATE/TIME OF TEST: 01/13/2019 16:29:02

Diagnostics Check		OK	16:26	
Air Blank		0.000	16:26	
Control Test		0.081	16:26	
Air Blank		0.000	16:27	
Subject Sample	# 1		16:27	Subject Test Refused
Breath Volume		0.000	16:27	
Air Blank		0.000	16:28	
Control Test		0.080	16:28	
Air Blank		0.000	16:28	
Diagnostics Check		OK	16:29	

Data Download Date/Time: 01/22/2019 13:47

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAGE,D,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:08

SUBJECT'S LAST NAME: GAMBER

DATE/TIME OF TEST: 01/13/2019 20:34:47

Diagnostics Check		OK	20:28
Air Blank		0.000	20:29
Control Test		0.080	20:29
Air Blank		0.000	20:29
Subject Sample	# 1	0.188	20:30
Breath Volume		2.335	20:30
Air Blank		0.000	20:30
Air Blank		0.000	20:32
Subject Sample	# 2	0.189	20:33
Breath Volume		3.335	20:33
Air Blank		0.000	20:33
Control Test		0.080	20:34
Air Blank		0.000	20:34
Diagnostics Check		OK	20:34

Data Download Date/Time: 01/22/2019 13:47

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAGE,D,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HARPIN,R
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: GREE

DATE/TIME OF TEST: 01/13/2019 22:01:51

Diagnostics Check		OK	21:58	
Air Blank		0.000	21:59	
Control Test		0.080	21:59	
Air Blank		0.000	22:00	
Subject Sample	# 1		22:00	Subject Test Refused
Breath Volume		0.000	22:00	
Air Blank		0.000	22:00	
Control Test		0.080	22:01	
Air Blank		0.000	22:01	
Diagnostics Check		OK	22:01	

Data Download Date/Time: 01/22/2019 13:47

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: IVEY,P
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: SANDERS

DATE/TIME OF TEST: 01/18/2019 22:35:05

Diagnostics Check		OK	22:28
Air Blank		0.000	22:28
Control Test		0.081	22:29
Air Blank		0.000	22:29
Subject Sample	# 1	0.000	22:30
Breath Volume		1.589	22:30
Air Blank		0.000	22:30
Air Blank		0.000	22:32
Subject Sample	# 2	0.000	22:33
Breath Volume		1.910	22:33
Air Blank		0.000	22:34
Control Test		0.081	22:34
Air Blank		0.000	22:35
Diagnostics Check		OK	22:35

Data Download Date/Time: 01/22/2019 13:47

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 01/20/2019 02:22:08

Diagnostics Check		OK	02:12
Air Blank		0.000	02:12
Control Test		0.080	02:12
Air Blank		0.000	02:13
Subject Sample	# 1	0.129	02:14
Breath Volume		2.949	02:14
Air Blank		0.000	02:14
Air Blank		0.000	02:16
Subject Sample	# 2	0.107	02:17
Breath Volume		1.398	02:17
Air Blank		0.000	02:18
Air Blank		0.000	02:20
Subject Sample	# 3	0.118	02:20
Breath Volume		2.730	02:20
Air Blank		0.000	02:21
Control Test		0.080	02:21
Air Blank		0.000	02:22
Diagnostics Check		OK	02:22

Data Download Date/Time: 01/22/2019 13:47

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: COHILL,D
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: KRAKOWIAK

DATE/TIME OF TEST: 01/20/2019 04:37:25

Diagnostics Check		OK	04:30
Air Blank		0.000	04:31
Control Test		0.080	04:31
Air Blank		0.000	04:32
Subject Sample	# 1	0.163	04:32
Breath Volume		2.687	04:32
Air Blank		0.000	04:33
Air Blank		0.000	04:35
Subject Sample	# 2	0.161	04:35
Breath Volume		3.246	04:35
Air Blank		0.000	04:36
Control Test		0.080	04:36
Air Blank		0.000	04:37
Diagnostics Check		OK	04:37

Data Download Date/Time: 01/22/2019 13:47

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 01/21/2019 01:12:57

Diagnostics Check		OK	01:06
Air Blank		0.000	01:07
Control Test		0.081	01:07
Air Blank		0.000	01:07
Subject Sample	# 1	0.116	01:08
Breath Volume		1.289	01:08
Air Blank		0.000	01:08
Air Blank		0.000	01:10
Subject Sample	# 2	0.119	01:11
Breath Volume		1.644	01:11
Air Blank		0.000	01:12
Control Test		0.081	01:12
Air Blank		0.000	01:12
Diagnostics Check		OK	01:12

Data Download Date/Time: 01/22/2019 13:47

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: DEAN

DATE/TIME OF TEST: 01/21/2019 02:59:35

Diagnostics Check		OK	02:56	
Air Blank		0.000	02:57	
Control Test		0.081	02:57	
Air Blank		0.000	02:58	
Subject Sample	# 1		02:58	Subject Test Refused
Breath Volume		0.000	02:58	
Air Blank		0.000	02:58	
Control Test		0.082	02:59	
Air Blank		0.000	02:59	
Diagnostics Check		OK	02:59	

Data Download Date/Time: 01/22/2019 13:47

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ZINZ,P
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:47

SUBJECT'S LAST NAME: SNELL

DATE/TIME OF TEST: 01/21/2019 05:13:52

Diagnostics Check		OK	05:07
Air Blank		0.000	05:08
Control Test		0.081	05:08
Air Blank		0.000	05:08
Subject Sample	# 1	0.232	05:09
Breath Volume		1.972	05:09
Air Blank		0.000	05:10
Air Blank		0.000	05:11
Subject Sample	# 2	0.225	05:12
Breath Volume		1.921	05:12
Air Blank		0.000	05:12
Control Test		0.081	05:13
Air Blank		0.000	05:13
Diagnostics Check		OK	05:13

Data Download Date/Time: 01/22/2019 13:47

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: MOUKAKOS

DATE/TIME OF TEST: 01/22/2019 03:44:47

Diagnostics Check		OK	03:37
Air Blank		0.000	03:38
Control Test		0.081	03:38
Air Blank		0.000	03:39
Subject Sample	# 1	0.142	03:40
Breath Volume		1.683	03:40
Air Blank		0.000	03:40
Air Blank		0.000	03:42
Subject Sample	# 2	0.133	03:43
Breath Volume		1.316	03:43
Air Blank		0.000	03:43
Control Test		0.080	03:44
Air Blank		0.000	03:44
Diagnostics Check		OK	03:44

Data Download Date/Time: 01/22/2019 13:47

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLDEN,E,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: HIGGINS

DATE/TIME OF TEST: 01/25/2019 21:54:24

Diagnostics Check		OK	21:51	
Air Blank		0.000	21:51	
Control Test		0.081	21:51	
Air Blank		0.000	21:52	
Subject Sample	# 1		21:53	Subject Test Refused
Breath Volume		0.000	21:53	
Air Blank		0.000	21:53	
Control Test		0.082	21:53	
Air Blank		0.000	21:54	
Diagnostics Check		OK	21:54	

Data Download Date/Time: 02/07/2019 12:58

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLDEN,E,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ALBERT,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: LUMPKIN

DATE/TIME OF TEST: 01/25/2019 22:43:48

Diagnostics Check		OK	22:37
Air Blank		0.000	22:37
Control Test		0.082	22:38
Air Blank		0.000	22:38
Subject Sample	# 1	0.000	22:39
Breath Volume		3.523	22:39
Air Blank		0.000	22:39
Air Blank		0.000	22:41
Subject Sample	# 2	0.000	22:42
Breath Volume		3.691	22:42
Air Blank		0.000	22:42
Control Test		0.081	22:43
Air Blank		0.000	22:43
Diagnostics Check		OK	22:43

Data Download Date/Time: 02/07/2019 12:58

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 01/25/2019 23:45:52

Diagnostics Check		OK	23:39
Air Blank		0.000	23:40
Control Test		0.082	23:40
Air Blank		0.000	23:41
Subject Sample	# 1	0.151	23:41
Breath Volume		1.972	23:41
Air Blank		0.000	23:42
Air Blank		0.000	23:44
Subject Sample	# 2	0.148	23:44
Breath Volume		1.753	23:44
Air Blank		0.000	23:44
Control Test		0.081	23:45
Air Blank		0.000	23:45
Diagnostics Check		OK	23:45

Data Download Date/Time: 02/07/2019 12:58

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNS,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: STANG

DATE/TIME OF TEST: 01/26/2019 02:48:36

Diagnostics Check		OK	02:42
Air Blank		0.000	02:42
Control Test		0.081	02:43
Air Blank		0.000	02:43
Subject Sample	# 1	0.172	02:44
Breath Volume		1.953	02:44
Air Blank		0.000	02:44
Air Blank		0.000	02:46
Subject Sample	# 2	0.176	02:47
Breath Volume		1.898	02:47
Air Blank		0.000	02:47
Control Test		0.082	02:48
Air Blank		0.000	02:48
Diagnostics Check		OK	02:48

Data Download Date/Time: 02/07/2019 12:58

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HAWKINS,A
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: KIM

DATE/TIME OF TEST: 01/26/2019 04:07:13

Diagnostics Check		OK	04:04	
Air Blank		0.000	04:04	
Control Test		0.081	04:05	
Air Blank		0.000	04:05	
Subject Sample	# 1		04:05	Subject Test Refused
Breath Volume		0.000	04:05	
Air Blank		0.000	04:06	
Control Test		0.082	04:06	
Air Blank		0.000	04:07	
Diagnostics Check		OK	04:07	

Data Download Date/Time: 02/07/2019 12:58

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: COHILL,D
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: COYKENDALL

DATE/TIME OF TEST: 01/26/2019 04:50:09

Diagnostics Check		OK	04:43
Air Blank		0.000	04:44
Control Test		0.082	04:44
Air Blank		0.000	04:45
Subject Sample	# 1	0.163	04:45
Breath Volume		2.351	04:45
Air Blank		0.000	04:46
Air Blank		0.000	04:47
Subject Sample	# 2	0.156	04:48
Breath Volume		2.574	04:48
Air Blank		0.000	04:49
Control Test		0.082	04:49
Air Blank		0.000	04:50
Diagnostics Check		OK	04:50

Data Download Date/Time: 02/07/2019 12:58

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: GRANVILLE

DATE/TIME OF TEST: 01/26/2019 23:16:15

Diagnostics Check		OK	23:13	
Air Blank		0.000	23:13	
Control Test		0.081	23:14	
Air Blank		0.000	23:14	
Subject Sample	# 1		23:14	Subject Test Refused
Breath Volume		0.000	23:14	
Air Blank		0.000	23:15	
Control Test		0.082	23:15	
Air Blank		0.000	23:16	
Diagnostics Check		OK	23:16	

Data Download Date/Time: 02/07/2019 12:58

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: MEZA URROZ

DATE/TIME OF TEST: 01/27/2019 01:27:46

Diagnostics Check		OK	01:21
Air Blank		0.000	01:22
Control Test		0.081	01:22
Air Blank		0.000	01:23
Subject Sample	# 1	0.165	01:23
Breath Volume		2.343	01:23
Air Blank		0.000	01:24
Air Blank		0.000	01:25
Subject Sample	# 2	0.163	01:26
Breath Volume		2.609	01:26
Air Blank		0.000	01:26
Control Test		0.081	01:27
Air Blank		0.000	01:27
Diagnostics Check		OK	01:27

Data Download Date/Time: 02/07/2019 12:58

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNS,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: BARRS

DATE/TIME OF TEST: 01/27/2019 23:43:51

Diagnostics Check		OK	23:41	
Air Blank		0.000	23:41	
Control Test		0.081	23:41	
Air Blank		0.000	23:42	
Subject Sample	# 1		23:42	Subject Test Refused
Breath Volume		0.000	23:42	
Air Blank		0.000	23:42	
Control Test		0.081	23:43	
Air Blank		0.000	23:43	
Diagnostics Check		OK	23:43	

Data Download Date/Time: 02/07/2019 12:58

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BUTLER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: MYRICK

DATE/TIME OF TEST: 01/28/2019 00:52:00

Diagnostics Check		OK	00:44
Air Blank		0.000	00:45
Control Test		0.081	00:45
Air Blank		0.000	00:46
Subject Sample	# 1	0.157	00:47
Breath Volume		1.675	00:47
Air Blank		0.000	00:48
Air Blank		0.000	00:50
Subject Sample	# 2	0.170	00:50
Breath Volume		1.191	00:50
Air Blank		0.000	00:51
Control Test		0.080	00:51
Air Blank		0.000	00:51
Diagnostics Check		OK	00:52

Data Download Date/Time: 02/07/2019 12:58

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: MERCEDEZ DUNLAP

DATE/TIME OF TEST: 01/28/2019 01:56:33

Diagnostics Check		OK	01:53	
Air Blank		0.000	01:54	
Control Test		0.081	01:54	
Air Blank		0.000	01:55	
Subject Sample	# 1		01:55	Subject Test Refused
Breath Volume		0.000	01:55	
Air Blank		0.000	01:55	
Control Test		0.081	01:55	
Air Blank		0.000	01:56	
Diagnostics Check		OK	01:56	

Data Download Date/Time: 02/07/2019 12:58

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HAWKINS,A
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:57

SUBJECT'S LAST NAME: ENCARNACION RUBIERA

DATE/TIME OF TEST: 01/28/2019 05:23:18

Diagnostics Check		OK	05:17
Air Blank		0.000	05:17
Control Test		0.081	05:17
Air Blank		0.000	05:18
Subject Sample	# 1	0.132	05:18
Breath Volume		2.851	05:18
Air Blank		0.000	05:19
Air Blank		0.000	05:21
Subject Sample	# 2	0.139	05:21
Breath Volume		1.976	05:21
Air Blank		0.000	05:22
Control Test		0.082	05:22
Air Blank		0.000	05:23
Diagnostics Check		OK	05:23

Data Download Date/Time: 02/07/2019 12:58

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: EVERTON

DATE/TIME OF TEST: 01/29/2019 00:20:38

Diagnostics Check		OK	00:14
Air Blank		0.000	00:14
Control Test		0.080	00:15
Air Blank		0.000	00:15
Subject Sample	# 1	0.150	00:16
Breath Volume		3.042	00:16
Air Blank		0.000	00:16
Air Blank		0.000	00:18
Subject Sample	# 2	0.147	00:19
Breath Volume		2.339	00:19
Air Blank		0.000	00:19
Control Test		0.080	00:20
Air Blank		0.000	00:20
Diagnostics Check		OK	00:20

Data Download Date/Time: 02/07/2019 12:58

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PATTERSON,K
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: SAAVEDRA

DATE/TIME OF TEST: 01/29/2019 02:56:50

Diagnostics Check		OK	02:53	
Air Blank		0.000	02:54	
Control Test		0.081	02:54	
Air Blank		0.000	02:55	
Subject Sample	# 1		02:55	Subject Test Refused
Breath Volume		0.000	02:55	
Air Blank		0.000	02:55	
Control Test		0.081	02:56	
Air Blank		0.000	02:56	
Diagnostics Check		OK	02:56	

Data Download Date/Time: 02/07/2019 12:58

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SORRELLS,M
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: WYATT

DATE/TIME OF TEST: 01/29/2019 03:45:44

Diagnostics Check		OK	03:39
Air Blank		0.000	03:40
Control Test		0.081	03:40
Air Blank		0.000	03:40
Subject Sample	# 1	0.177	03:41
Breath Volume		2.527	03:41
Air Blank		0.000	03:42
Air Blank		0.000	03:43
Subject Sample	# 2	0.177	03:44
Breath Volume		2.835	03:44
Air Blank		0.000	03:44
Control Test		0.080	03:45
Air Blank		0.000	03:45
Diagnostics Check		OK	03:45

Data Download Date/Time: 02/07/2019 12:58

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PERRY,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:54

SUBJECT'S LAST NAME: SUBEK

DATE/TIME OF TEST: 01/29/2019 05:20:36

Diagnostics Check		OK	05:17	
Air Blank		0.000	05:18	
Control Test		0.081	05:18	
Air Blank		0.000	05:19	
Subject Sample	# 1		05:19	Subject Test Refused
Breath Volume		0.000	05:19	
Air Blank		0.000	05:19	
Control Test		0.080	05:20	
Air Blank		0.000	05:20	
Diagnostics Check		OK	05:20	

Data Download Date/Time: 02/07/2019 12:58

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARIDA,K,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SPENCER,K
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 06:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:27

SUBJECT'S LAST NAME: PARKS

DATE/TIME OF TEST: 01/29/2019 08:53:48

Diagnostics Check		OK	08:47
Air Blank		0.000	08:47
Control Test		0.081	08:48
Air Blank		0.000	08:48
Subject Sample	# 1	0.144	08:49
Breath Volume		4.355	08:49
Air Blank		0.000	08:49
Air Blank		0.000	08:51
Subject Sample	# 2	0.149	08:52
Breath Volume		4.113	08:52
Air Blank		0.000	08:52
Control Test		0.081	08:53
Air Blank		0.000	08:53
Diagnostics Check		OK	08:53

Data Download Date/Time: 02/07/2019 12:58

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JENSEN,D
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: BECK

DATE/TIME OF TEST: 01/30/2019 00:02:29

Diagnostics Check		OK	23:59	
Air Blank		0.000	00:00	
Control Test		0.081	00:00	
Air Blank		0.000	00:00	
Subject Sample	# 1		00:01	Subject Test Refused
Breath Volume		0.000	00:01	
Air Blank		0.000	00:01	
Control Test		0.081	00:01	
Air Blank		0.000	00:02	
Diagnostics Check		OK	00:02	

Data Download Date/Time: 02/07/2019 12:58

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PATTERSON,K
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: SANDERS

DATE/TIME OF TEST: 01/30/2019 04:13:02

Diagnostics Check		OK	04:07
Air Blank		0.000	04:07
Control Test		0.081	04:08
Air Blank		0.000	04:08
Subject Sample	# 1	0.000	04:08
Breath Volume		2.839	04:08
Air Blank		0.000	04:09
Air Blank		0.000	04:11
Subject Sample	# 2	0.000	04:11
Breath Volume		3.406	04:11
Air Blank		0.000	04:12
Control Test		0.082	04:12
Air Blank		0.000	04:12
Diagnostics Check		OK	04:13

Data Download Date/Time: 02/07/2019 12:59

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: NEWMAN

DATE/TIME OF TEST: 01/31/2019 01:49:56

Diagnostics Check		OK	01:43
Air Blank		0.000	01:44
Control Test		0.082	01:44
Air Blank		0.000	01:45
Subject Sample	# 1	0.114	01:45
Breath Volume		3.472	01:45
Air Blank		0.000	01:46
Air Blank		0.000	01:48
Subject Sample	# 2	0.105	01:48
Breath Volume		3.425	01:48
Air Blank		0.000	01:49
Control Test		0.081	01:49
Air Blank		0.000	01:49
Diagnostics Check		OK	01:49

Data Download Date/Time: 02/07/2019 12:59

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 01/31/2019 03:32:59

Diagnostics Check		OK	03:30	
Air Blank		0.000	03:30	
Control Test		0.082	03:30	
Air Blank		0.000	03:31	
Subject Sample	# 1		03:31	Subject Test Refused
Breath Volume		0.000	03:31	
Air Blank		0.000	03:32	
Control Test		0.082	03:32	
Air Blank		0.000	03:32	
Diagnostics Check		OK	03:32	

Data Download Date/Time: 02/07/2019 12:59

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOHN,ANTHONY,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: HAMILTON

DATE/TIME OF TEST: 01/31/2019 22:20:33

Diagnostics Check		OK	22:14
Air Blank		0.000	22:14
Control Test		0.081	22:15
Air Blank		0.000	22:15
Subject Sample	# 1	0.172	22:16
Breath Volume		2.339	22:16
Air Blank		0.000	22:16
Air Blank		0.000	22:18
Subject Sample	# 2	0.164	22:19
Breath Volume		1.648	22:19
Air Blank		0.000	22:19
Control Test		0.080	22:19
Air Blank		0.000	22:20
Diagnostics Check		OK	22:20

Data Download Date/Time: 02/07/2019 12:59

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JONES,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 01/01/2019 03:11:43

Diagnostics Check		OK	03:08	
Air Blank		0.000	03:09	
Control Test		0.080	03:09	
Air Blank		0.000	03:10	
Subject Sample	# 1		03:10	Subject Test Refused
Breath Volume		0.000	03:10	
Air Blank		0.000	03:10	
Control Test		0.080	03:11	
Air Blank		0.000	03:11	
Diagnostics Check		OK	03:11	

Data Download Date/Time: 01/22/2019 13:32

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: DOMINGUES DEL CID

DATE/TIME OF TEST: 01/01/2019 05:04:51

Diagnostics Check		OK	04:58
Air Blank		0.000	04:59
Control Test		0.080	04:59
Air Blank		0.000	05:00
Subject Sample	# 1	0.086	05:00
Breath Volume		2.375	05:00
Air Blank		0.000	05:01
Air Blank		0.000	05:03
Subject Sample	# 2	0.087	05:03
Breath Volume		2.671	05:03
Air Blank		0.000	05:03
Control Test		0.080	05:04
Air Blank		0.000	05:04
Diagnostics Check		OK	05:04

Data Download Date/Time: 01/22/2019 13:32

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNS,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:17

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 01/01/2019 05:40:48

Diagnostics Check		OK	05:37	
Air Blank		0.000	05:38	
Control Test		0.080	05:38	
Air Blank		0.000	05:39	
Subject Sample	# 1		05:39	Subject Test Refused
Breath Volume		0.000	05:39	
Air Blank		0.000	05:39	
Control Test		0.080	05:40	
Air Blank		0.000	05:40	
Diagnostics Check		OK	05:40	

Data Download Date/Time: 01/22/2019 13:32

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMAS,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: CHANNELL

DATE/TIME OF TEST: 01/02/2019 01:58:12

Diagnostics Check		OK	01:52
Air Blank		0.000	01:52
Control Test		0.080	01:52
Air Blank		0.000	01:53
Subject Sample	# 1	0.000	01:53
Breath Volume		1.746	01:53
Air Blank		0.000	01:54
Air Blank		0.000	01:56
Subject Sample	# 2	0.000	01:56
Breath Volume		1.546	01:56
Air Blank		0.000	01:57
Control Test		0.080	01:57
Air Blank		0.000	01:58
Diagnostics Check		OK	01:58

Data Download Date/Time: 01/22/2019 13:32

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MARRERO,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: ORTIZ DOMINGO

DATE/TIME OF TEST: 01/02/2019 02:44:12

Diagnostics Check		OK	02:37
Air Blank		0.000	02:38
Control Test		0.080	02:38
Air Blank		0.000	02:38
Subject Sample	# 1	0.220	02:39
Breath Volume		3.289	02:39
Air Blank		0.000	02:40
Air Blank		0.000	02:42
Subject Sample	# 2	0.221	02:42
Breath Volume		2.980	02:42
Air Blank		0.000	02:43
Control Test		0.080	02:43
Air Blank		0.000	02:44
Diagnostics Check		OK	02:44

Data Download Date/Time: 01/22/2019 13:32

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONNER,CHARLES,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRYANT,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: QUARLES

DATE/TIME OF TEST: 01/02/2019 22:47:52

Diagnostics Check		OK	22:41
Air Blank		0.000	22:41
Control Test		0.080	22:42
Air Blank		0.000	22:42
Subject Sample	# 1	0.201	22:43
Breath Volume		1.753	22:43
Air Blank		0.000	22:43
Air Blank		0.000	22:45
Subject Sample	# 2	0.216	22:46
Breath Volume		2.312	22:46
Air Blank		0.000	22:46
Control Test		0.079	22:47
Air Blank		0.000	22:47
Diagnostics Check		OK	22:47

Data Download Date/Time: 01/22/2019 13:32

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRYANT,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: ROMO DE ALBA

DATE/TIME OF TEST: 01/03/2019 22:57:42

Diagnostics Check		OK	22:50
Air Blank		0.000	22:50
Control Test		0.080	22:51
Air Blank		0.000	22:51
Subject Sample	# 1	0.188	22:52
Breath Volume		1.664	22:52
Air Blank		0.000	22:53
Air Blank		0.000	22:55
Subject Sample	# 2	0.186	22:56
Breath Volume		1.378	22:56
Air Blank		0.000	22:56
Control Test		0.079	22:57
Air Blank		0.000	22:57
Diagnostics Check		OK	22:57

Data Download Date/Time: 01/22/2019 13:33

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: EPHRON

DATE/TIME OF TEST: 01/03/2019 23:53:59

Diagnostics Check		OK	23:48	
Air Blank		0.000	23:48	
Control Test		0.080	23:48	
Air Blank		0.000	23:49	
Subject Sample	# 1	0.146	23:50	
Breath Volume		1.226	23:50	
Air Blank		0.000	23:51	
Air Blank		0.000	23:53	
Subject Sample	# 2		23:53	RFI Detect
Breath Volume		0.453	23:53	
Air Blank			23:53	RFI Detect
Air Blank		0.000	23:53	

Data Download Date/Time: 01/22/2019 13:33

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: EPHRON

DATE/TIME OF TEST: 01/04/2019 00:02:33

Diagnostics Check		OK	23:55
Air Blank		0.000	23:56
Control Test		0.079	23:56
Air Blank		0.000	23:57
Subject Sample	# 2	0.142	23:57
Breath Volume		1.253	23:57
Air Blank		0.000	23:58
Air Blank		0.000	00:00
Subject Sample	# 1	0.143	00:01
Breath Volume		1.476	00:01
Air Blank		0.000	00:01
Control Test		0.079	00:01
Air Blank		0.000	00:02
Diagnostics Check		OK	00:02

Data Download Date/Time: 01/22/2019 13:33

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: ELLIS

DATE/TIME OF TEST: 01/04/2019 03:50:09

Diagnostics Check		OK	03:43
Air Blank		0.000	03:44
Control Test		0.079	03:44
Air Blank		0.000	03:45
Subject Sample	# 1	0.079	03:45
Breath Volume		2.050	03:45
Air Blank		0.000	03:46
Air Blank		0.000	03:48
Subject Sample	# 2	0.083	03:48
Breath Volume		2.121	03:48
Air Blank		0.000	03:49
Control Test		0.081	03:49
Air Blank		0.000	03:50
Diagnostics Check		OK	03:50

Data Download Date/Time: 01/22/2019 13:33

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: WOOD

DATE/TIME OF TEST: 01/05/2019 00:57:07

Diagnostics Check		OK	00:54	
Air Blank		0.000	00:54	
Control Test		0.080	00:54	
Air Blank		0.000	00:55	
Subject Sample	# 1		00:55	Subject Test Refused
Breath Volume		0.000	00:55	
Air Blank		0.000	00:56	
Control Test		0.080	00:56	
Air Blank		0.000	00:57	
Diagnostics Check		OK	00:57	

Data Download Date/Time: 01/22/2019 13:33

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BEALL,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: ROSSETTI

DATE/TIME OF TEST: 01/05/2019 03:59:00

Diagnostics Check		OK	03:55	
Air Blank		0.000	03:55	
Control Test		0.081	03:56	
Air Blank		0.000	03:56	
Subject Sample	# 1		03:57	Subject Test Refused
Breath Volume		0.000	03:57	
Air Blank		0.000	03:58	
Control Test		0.080	03:58	
Air Blank		0.000	03:58	
Diagnostics Check		OK	03:59	

Data Download Date/Time: 01/22/2019 13:33

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: SIMIC

DATE/TIME OF TEST: 01/06/2019 01:07:48

Diagnostics Check		OK	01:01
Air Blank		0.000	01:02
Control Test		0.081	01:02
Air Blank		0.000	01:03
Subject Sample	# 1	0.099	01:03
Breath Volume		2.875	01:03
Air Blank		0.000	01:04
Air Blank		0.000	01:05
Subject Sample	# 2	0.095	01:06
Breath Volume		2.429	01:06
Air Blank		0.000	01:06
Control Test		0.080	01:07
Air Blank		0.000	01:07
Diagnostics Check		OK	01:07

Data Download Date/Time: 01/22/2019 13:33

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNSON,G
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 01/06/2019 04:08:19

Diagnostics Check		OK	04:02
Air Blank		0.000	04:02
Control Test		0.081	04:02
Air Blank		0.000	04:03
Subject Sample	# 1	0.119	04:03
Breath Volume		3.421	04:03
Air Blank		0.000	04:04
Air Blank		0.000	04:06
Subject Sample	# 2	0.119	04:06
Breath Volume		3.222	04:06
Air Blank		0.000	04:07
Control Test		0.080	04:07
Air Blank		0.000	04:08
Diagnostics Check		OK	04:08

Data Download Date/Time: 01/22/2019 13:33

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: CARRASQUILLO

DATE/TIME OF TEST: 01/06/2019 05:09:11

Diagnostics Check		OK	05:03
Air Blank		0.000	05:03
Control Test		0.081	05:03
Air Blank		0.000	05:04
Subject Sample	# 1	0.166	05:04
Breath Volume		3.835	05:04
Air Blank		0.000	05:05
Air Blank		0.000	05:07
Subject Sample	# 2	0.158	05:07
Breath Volume		2.421	05:07
Air Blank		0.000	05:08
Control Test		0.080	05:08
Air Blank		0.000	05:09
Diagnostics Check		OK	05:09

Data Download Date/Time: 01/22/2019 13:33

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: MOZES

DATE/TIME OF TEST: 01/06/2019 23:38:53

Diagnostics Check		OK	23:36	
Air Blank		0.000	23:36	
Control Test		0.081	23:36	
Air Blank		0.000	23:37	
Subject Sample	# 1		23:37	Subject Test Refused
Breath Volume		0.000	23:37	
Air Blank		0.000	23:37	
Control Test		0.080	23:38	
Air Blank		0.000	23:38	
Diagnostics Check		OK	23:38	

Data Download Date/Time: 01/22/2019 13:33

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,F,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LANGDON,N
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 07:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:50

SUBJECT'S LAST NAME: ARCE

DATE/TIME OF TEST: 01/07/2019 10:18:33

Diagnostics Check		OK	10:12
Air Blank		0.000	10:12
Control Test		0.080	10:13
Air Blank		0.000	10:13
Subject Sample	# 1	0.000	10:14
Breath Volume		1.800	10:14
Air Blank		0.000	10:14
Air Blank		0.000	10:16
Subject Sample	# 2	0.000	10:17
Breath Volume		1.667	10:17
Air Blank		0.000	10:17
Control Test		0.080	10:17
Air Blank		0.000	10:18
Diagnostics Check		OK	10:18

Data Download Date/Time: 01/22/2019 13:33

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: DIXON

DATE/TIME OF TEST: 01/08/2019 00:55:43

Diagnostics Check		OK	00:49
Air Blank		0.000	00:50
Control Test		0.081	00:50
Air Blank		0.000	00:51
Subject Sample	# 1	0.000	00:51
Breath Volume		4.371	00:51
Air Blank		0.000	00:51
Air Blank		0.000	00:53
Subject Sample	# 2	0.000	00:54
Breath Volume		3.382	00:54
Air Blank		0.000	00:54
Control Test		0.081	00:55
Air Blank		0.000	00:55
Diagnostics Check		OK	00:55

Data Download Date/Time: 01/22/2019 13:33

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: CALLAHAN

DATE/TIME OF TEST: 01/08/2019 03:52:37

Diagnostics Check		OK	03:46
Air Blank		0.000	03:47
Control Test		0.081	03:47
Air Blank		0.000	03:47
Subject Sample	# 1	0.000	03:48
Breath Volume		2.558	03:48
Air Blank		0.000	03:48
Air Blank		0.000	03:50
Subject Sample	# 2	0.000	03:51
Breath Volume		2.554	03:51
Air Blank		0.000	03:51
Control Test		0.081	03:52
Air Blank		0.000	03:52
Diagnostics Check		OK	03:52

Data Download Date/Time: 01/22/2019 13:33

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: HESTER

DATE/TIME OF TEST: 01/09/2019 02:26:50

Diagnostics Check		OK	02:24	
Air Blank		0.000	02:24	
Control Test		0.081	02:24	
Air Blank		0.000	02:25	
Subject Sample	# 1		02:25	Subject Test Refused
Breath Volume		0.000	02:25	
Air Blank		0.000	02:25	
Control Test		0.080	02:26	
Air Blank		0.000	02:26	
Diagnostics Check		OK	02:26	

Data Download Date/Time: 01/22/2019 13:33

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRYANT,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: BELK

DATE/TIME OF TEST: 01/11/2019 23:22:05

Diagnostics Check		OK	23:15
Air Blank		0.000	23:15
Control Test		0.081	23:16
Air Blank		0.000	23:16
Subject Sample	# 1	0.294	23:17
Breath Volume		2.011	23:17
Air Blank		0.000	23:18
Air Blank		0.000	23:20
Subject Sample	# 2	0.292	23:20
Breath Volume		1.968	23:20
Air Blank		0.000	23:21
Control Test		0.080	23:21
Air Blank		0.000	23:22
Diagnostics Check		OK	23:22

Data Download Date/Time: 01/22/2019 13:33

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 01/12/2019 02:59:58

Diagnostics Check		OK	02:57	
Air Blank		0.000	02:57	
Control Test		0.081	02:57	
Air Blank		0.000	02:58	
Subject Sample	# 1		02:58	Subject Test Refused
Breath Volume		0.000	02:58	
Air Blank		0.000	02:59	
Control Test		0.081	02:59	
Air Blank		0.000	02:59	
Diagnostics Check		OK	02:59	

Data Download Date/Time: 01/22/2019 13:33

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,J,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: WEST

DATE/TIME OF TEST: 01/12/2019 23:52:28

Diagnostics Check		OK	23:49	
Air Blank		0.000	23:50	
Control Test		0.081	23:50	
Air Blank		0.000	23:50	
Subject Sample	# 1		23:51	Subject Test Refused
Breath Volume		0.000	23:51	
Air Blank		0.000	23:51	
Control Test		0.080	23:51	
Air Blank		0.000	23:52	
Diagnostics Check		OK	23:52	

Data Download Date/Time: 01/22/2019 13:33

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HYSMITH,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: AWI

DATE/TIME OF TEST: 01/13/2019 01:22:00

Diagnostics Check		OK	01:12
Air Blank		0.000	01:13
Control Test		0.080	01:13
Air Blank		0.000	01:13
Subject Sample	# 1	0.322	01:14
Breath Volume		3.234	01:14
Air Blank		0.000	01:15
Air Blank		0.000	01:17
Subject Sample	# 2	0.289	01:17
Breath Volume		1.933	01:17
Air Blank		0.000	01:18
Air Blank		0.000	01:20
Subject Sample	# 3	0.322	01:20
Breath Volume		3.390	01:20
Air Blank		0.000	01:21
Control Test		0.080	01:21
Air Blank		0.000	01:21
Diagnostics Check		OK	01:22

Data Download Date/Time: 01/22/2019 13:33

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FILBERT,E
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: KIKER-BEURY

DATE/TIME OF TEST: 01/13/2019 01:58:03

Diagnostics Check	OK	01:56	
Air Blank	0.000	01:57	
Control Test	0.080	01:57	
Air Blank		01:57	RFI Detect
Air Blank	0.000	01:58	

Data Download Date/Time: 01/22/2019 13:33

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FILBERT,E
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: KIKER-BEURY

DATE/TIME OF TEST: 01/13/2019 02:04:09

Diagnostics Check		OK	02:01	
Air Blank		0.000	02:01	
Control Test		0.080	02:02	
Air Blank		0.000	02:02	
Subject Sample	# 1		02:02	Subject Test Refused
Breath Volume		0.000	02:02	
Air Blank		0.000	02:03	
Control Test		0.080	02:03	
Air Blank		0.000	02:04	
Diagnostics Check		OK	02:04	

Data Download Date/Time: 01/22/2019 13:33

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: NUNES

DATE/TIME OF TEST: 01/13/2019 03:43:56

Diagnostics Check		OK	03:37
Air Blank		0.000	03:38
Control Test		0.081	03:38
Air Blank		0.000	03:39
Subject Sample	# 1	0.100	03:39
Breath Volume		1.238	03:39
Air Blank		0.000	03:39
Air Blank		0.000	03:41
Subject Sample	# 2	0.101	03:42
Breath Volume		1.710	03:42
Air Blank		0.000	03:43
Control Test		0.080	03:43
Air Blank		0.000	03:43
Diagnostics Check		OK	03:43

Data Download Date/Time: 01/22/2019 13:33

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JENSEN,D
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: ASHFORD

DATE/TIME OF TEST: 01/13/2019 04:28:37

Diagnostics Check		OK	04:24	
Air Blank		0.000	04:24	
Control Test		0.080	04:25	
Air Blank		0.000	04:25	
Subject Sample	# 1		04:27	Subject Test Refused
Breath Volume		0.000	04:27	
Air Blank		0.000	04:27	
Control Test		0.080	04:28	
Air Blank		0.000	04:28	
Diagnostics Check		OK	04:28	

Data Download Date/Time: 01/22/2019 13:33

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: FIRTH

DATE/TIME OF TEST: 01/16/2019 00:57:44

Diagnostics Check		OK	00:54	
Air Blank		0.000	00:55	
Control Test		0.081	00:55	
Air Blank		0.000	00:56	
Subject Sample	# 1		00:56	Subject Test Refused
Breath Volume		0.000	00:56	
Air Blank		0.000	00:56	
Control Test		0.081	00:57	
Air Blank		0.000	00:57	
Diagnostics Check		OK	00:57	

Data Download Date/Time: 01/22/2019 13:33

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: EAKINS

DATE/TIME OF TEST: 01/16/2019 02:58:45

Diagnostics Check		OK	02:51
Air Blank		0.000	02:51
Control Test		0.081	02:52
Air Blank		0.000	02:52
Subject Sample	# 1	0.094	02:54
Breath Volume		1.437	02:54
Air Blank		0.000	02:54
Air Blank		0.000	02:56
Subject Sample	# 2	0.102	02:57
Breath Volume		1.769	02:57
Air Blank		0.000	02:57
Control Test		0.081	02:58
Air Blank		0.000	02:58
Diagnostics Check		OK	02:58

Data Download Date/Time: 01/22/2019 13:33

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CAMACHO,K
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: STEADMAN

DATE/TIME OF TEST: 01/17/2019 01:23:37

Diagnostics Check		OK	01:20	
Air Blank		0.000	01:21	
Control Test		0.082	01:21	
Air Blank		0.000	01:22	
Subject Sample	# 1		01:22	Subject Test Refused
Breath Volume		0.000	01:22	
Air Blank		0.000	01:22	
Control Test		0.083	01:23	
Air Blank		0.000	01:23	
Diagnostics Check		OK	01:23	

Data Download Date/Time: 01/22/2019 13:33

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 01/18/2019 23:56:04

Diagnostics Check		OK	23:53	
Air Blank		0.000	23:53	
Control Test		0.081	23:54	
Air Blank		0.000	23:54	
Subject Sample	# 1		23:54	Subject Test Refused
Breath Volume		0.000	23:54	
Air Blank		0.000	23:55	
Control Test		0.081	23:55	
Air Blank		0.000	23:55	
Diagnostics Check		OK	23:56	

Data Download Date/Time: 01/22/2019 13:33

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: NEWMAN

DATE/TIME OF TEST: 01/19/2019 00:33:44

Diagnostics Check		OK	00:30	
Air Blank		0.000	00:31	
Control Test		0.081	00:31	
Air Blank		0.000	00:32	
Subject Sample	# 1		00:32	Subject Test Refused
Breath Volume		0.000	00:32	
Air Blank		0.000	00:32	
Control Test		0.081	00:33	
Air Blank		0.000	00:33	
Diagnostics Check		OK	00:33	

Data Download Date/Time: 01/22/2019 13:33

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: AUSTIN

DATE/TIME OF TEST: 01/19/2019 03:48:22

Diagnostics Check		OK	03:45	
Air Blank		0.000	03:45	
Control Test		0.081	03:46	
Air Blank		0.000	03:46	
Subject Sample	# 1		03:46	RFI Detect
Breath Volume		0.960	03:46	
Air Blank		0.000	03:47	
Control Test		0.081	03:47	
Air Blank		0.000	03:48	
Diagnostics Check		OK	03:48	

Data Download Date/Time: 01/22/2019 13:33

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: AUSTIN

DATE/TIME OF TEST: 01/19/2019 03:56:15

Diagnostics Check		OK	03:49
Air Blank		0.000	03:50
Control Test		0.081	03:50
Air Blank		0.000	03:50
Subject Sample	# 1	0.019	03:51
Breath Volume		1.125	03:51
Air Blank		0.000	03:52
Air Blank		0.000	03:54
Subject Sample	# 2	0.019	03:54
Breath Volume		2.160	03:54
Air Blank		0.000	03:55
Control Test		0.081	03:55
Air Blank		0.000	03:56
Diagnostics Check		OK	03:56

Data Download Date/Time: 01/22/2019 13:33

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIQUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: HINDMARSH

DATE/TIME OF TEST: 01/19/2019 04:48:19

Diagnostics Check		OK	04:42
Air Blank		0.000	04:42
Control Test		0.081	04:42
Air Blank		0.000	04:43
Subject Sample	# 1	0.264	04:43
Breath Volume		3.773	04:43
Air Blank		0.000	04:44
Air Blank		0.000	04:46
Subject Sample	# 2	0.274	04:46
Breath Volume		2.968	04:46
Air Blank		0.000	04:47
Control Test		0.081	04:47
Air Blank		0.000	04:48
Diagnostics Check		OK	04:48

Data Download Date/Time: 01/22/2019 13:33

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOINER,D
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: PURSELL

DATE/TIME OF TEST: 01/20/2019 02:21:07

Diagnostics Check		OK	02:18	
Air Blank		0.000	02:18	
Control Test		0.080	02:19	
Air Blank		0.000	02:19	
Subject Sample	# 1		02:19	Subject Test Refused
Breath Volume		0.000	02:19	
Air Blank		0.000	02:20	
Control Test		0.079	02:20	
Air Blank		0.000	02:21	
Diagnostics Check		OK	02:21	

Data Download Date/Time: 01/22/2019 13:33

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DRAGOMIR,I
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 01/20/2019 04:51:55

Diagnostics Check		OK	04:45
Air Blank		0.000	04:46
Control Test		0.080	04:46
Air Blank		0.000	04:46
Subject Sample	# 1	0.158	04:47
Breath Volume		4.300	04:47
Air Blank		0.000	04:48
Air Blank		0.000	04:49
Subject Sample	# 2	0.152	04:50
Breath Volume		3.757	04:50
Air Blank		0.000	04:51
Control Test		0.079	04:51
Air Blank		0.000	04:51
Diagnostics Check		OK	04:51

Data Download Date/Time: 01/22/2019 13:33

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 01/21/2019 01:36:31

Diagnostics Check		OK	01:30
Air Blank		0.000	01:31
Control Test		0.082	01:31
Air Blank		0.000	01:31
Subject Sample	# 1	0.000	01:32
Breath Volume		2.476	01:32
Air Blank		0.000	01:32
Air Blank		0.000	01:34
Subject Sample	# 2	0.000	01:35
Breath Volume		2.167	01:35
Air Blank		0.000	01:35
Control Test		0.082	01:35
Air Blank		0.000	01:36
Diagnostics Check		OK	01:36

Data Download Date/Time: 01/22/2019 13:33

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: COHILL,D
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: HOSNY

DATE/TIME OF TEST: 01/21/2019 04:14:54

Diagnostics Check		OK	04:08
Air Blank		0.000	04:08
Control Test		0.081	04:09
Air Blank		0.000	04:09
Subject Sample	# 1	0.147	04:10
Breath Volume		3.425	04:10
Air Blank		0.000	04:10
Air Blank		0.000	04:12
Subject Sample	# 2	0.145	04:13
Breath Volume		2.687	04:13
Air Blank		0.000	04:13
Control Test		0.082	04:14
Air Blank		0.000	04:14
Diagnostics Check		OK	04:14

Data Download Date/Time: 01/22/2019 13:33

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 01/22/2019 03:53:43

Diagnostics Check		OK	03:47
Air Blank		0.000	03:48
Control Test		0.081	03:48
Air Blank		0.000	03:48
Subject Sample	# 1	0.175	03:49
Breath Volume		1.644	03:49
Air Blank		0.000	03:50
Air Blank		0.000	03:51
Subject Sample	# 2	0.171	03:52
Breath Volume		1.589	03:52
Air Blank		0.000	03:52
Control Test		0.080	03:53
Air Blank		0.000	03:53
Diagnostics Check		OK	03:53

Data Download Date/Time: 01/22/2019 13:33

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 01/22/2019 23:34:37

Diagnostics Check		OK	23:28
Air Blank		0.000	23:28
Control Test		0.080	23:29
Air Blank		0.000	23:29
Subject Sample	# 1	0.236	23:30
Breath Volume		1.648	23:30
Air Blank		0.000	23:30
Air Blank		0.000	23:32
Subject Sample	# 2	0.233	23:33
Breath Volume		1.218	23:33
Air Blank		0.000	23:33
Control Test		0.080	23:34
Air Blank		0.000	23:34
Diagnostics Check		OK	23:34

Data Download Date/Time: 02/07/2019 12:53

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONNER,CHARLES,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:13

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 01/23/2019 21:40:14

Diagnostics Check		OK	21:33
Air Blank		0.000	21:33
Control Test		0.080	21:34
Air Blank		0.000	21:34
Subject Sample	# 1	0.153	21:35
Breath Volume		2.105	21:35
Air Blank		0.000	21:36
Air Blank		0.000	21:38
Subject Sample	# 2	0.151	21:38
Breath Volume		2.304	21:38
Air Blank		0.000	21:39
Control Test		0.080	21:39
Air Blank		0.000	21:40
Diagnostics Check		OK	21:40

Data Download Date/Time: 02/07/2019 12:53

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONNER,CHARLES,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: TENNANT

DATE/TIME OF TEST: 01/24/2019 20:34:33

Diagnostics Check		OK	20:31	
Air Blank		0.000	20:32	
Control Test		0.080	20:32	
Air Blank		0.000	20:33	
Subject Sample	# 1		20:33	Subject Test Refused
Breath Volume		0.000	20:33	
Air Blank		0.000	20:33	
Control Test		0.080	20:33	
Air Blank		0.000	20:34	
Diagnostics Check		OK	20:34	

Data Download Date/Time: 02/07/2019 12:53

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONNER,C,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: IVEY,P
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: MAGALLAN

DATE/TIME OF TEST: 01/24/2019 21:45:35

Diagnostics Check		OK	21:42	
Air Blank		0.000	21:42	
Control Test		0.081	21:43	
Air Blank		0.000	21:43	
Subject Sample	# 1		21:44	Subject Test Refused
Breath Volume		0.000	21:44	
Air Blank		0.000	21:44	
Control Test		0.080	21:44	
Air Blank		0.000	21:45	
Diagnostics Check		OK	21:45	

Data Download Date/Time: 02/07/2019 12:53

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: PAYNE

DATE/TIME OF TEST: 01/26/2019 02:44:01

Diagnostics Check		OK	02:37
Air Blank		0.000	02:37
Control Test		0.082	02:37
Air Blank		0.000	02:38
Subject Sample	# 1	0.096	02:39
Breath Volume		1.660	02:39
Air Blank		0.000	02:39
Air Blank		0.000	02:41
Subject Sample	# 2	0.094	02:42
Breath Volume		1.523	02:42
Air Blank		0.000	02:43
Control Test		0.082	02:43
Air Blank		0.000	02:43
Diagnostics Check		OK	02:44

Data Download Date/Time: 02/07/2019 12:53

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WALDREP,B
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 01/26/2019 03:29:32

Diagnostics Check		OK	03:26	
Air Blank		0.000	03:27	
Control Test		0.082	03:27	
Air Blank		0.000	03:28	
Subject Sample	# 1		03:28	Subject Test Refused
Breath Volume		0.000	03:28	
Air Blank		0.000	03:28	
Control Test		0.082	03:28	
Air Blank		0.000	03:29	
Diagnostics Check		OK	03:29	

Data Download Date/Time: 02/07/2019 12:53

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: TOMLIN

DATE/TIME OF TEST: 01/26/2019 04:42:11

Diagnostics Check		OK	04:39	
Air Blank		0.000	04:39	
Control Test		0.082	04:40	
Air Blank		0.000	04:40	
Subject Sample	# 1		04:40	Subject Test Refused
Breath Volume		0.000	04:40	
Air Blank		0.000	04:41	
Control Test		0.081	04:41	
Air Blank		0.000	04:42	
Diagnostics Check		OK	04:42	

Data Download Date/Time: 02/07/2019 12:53

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BARNES,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 06:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:04

SUBJECT'S LAST NAME: DIAZ VELAZQUEZ

DATE/TIME OF TEST: 01/26/2019 09:31:07

Diagnostics Check		OK	09:24
Air Blank		0.000	09:24
Control Test		0.082	09:25
Air Blank		0.000	09:25
Subject Sample	# 1	0.128	09:26
Breath Volume		1.945	09:26
Air Blank		0.000	09:27
Air Blank		0.000	09:29
Subject Sample	# 2	0.134	09:29
Breath Volume		2.195	09:29
Air Blank		0.000	09:30
Control Test		0.081	09:30
Air Blank		0.000	09:31
Diagnostics Check		OK	09:31

Data Download Date/Time: 02/07/2019 12:53

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DANZIE,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: WEST

DATE/TIME OF TEST: 01/27/2019 00:52:08

Diagnostics Check		OK	00:46	
Air Blank		0.000	00:47	
Control Test		0.081	00:47	
Air Blank		0.000	00:47	
Subject Sample	# 1	0.113	00:48	
Breath Volume		2.101	00:48	
Air Blank		0.000	00:48	
Air Blank		0.000	00:50	
Subject Sample	# 2	0.112	00:51	
Breath Volume		1.589	00:51	
Air Blank			00:51	RFI Detect
Air Blank		0.000	00:52	

Data Download Date/Time: 02/07/2019 12:53

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DANZIE,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: WEST

DATE/TIME OF TEST: 01/27/2019 00:59:55

Diagnostics Check		OK	00:53
Air Blank		0.000	00:54
Control Test		0.081	00:54
Air Blank		0.000	00:54
Subject Sample	# 1	0.110	00:55
Breath Volume		1.332	00:55
Air Blank		0.000	00:55
Air Blank		0.000	00:57
Subject Sample	# 2	0.113	00:58
Breath Volume		1.816	00:58
Air Blank		0.000	00:59
Control Test		0.081	00:59
Air Blank		0.000	00:59
Diagnostics Check		OK	00:59

Data Download Date/Time: 02/07/2019 12:53

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: COHILL,D
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: WEBER

DATE/TIME OF TEST: 01/27/2019 03:28:46

Diagnostics Check		OK	03:22
Air Blank		0.000	03:23
Control Test		0.081	03:23
Air Blank		0.000	03:23
Subject Sample	# 1	0.068	03:24
Breath Volume		3.023	03:24
Air Blank		0.000	03:24
Air Blank		0.000	03:26
Subject Sample	# 2	0.064	03:27
Breath Volume		1.781	03:27
Air Blank		0.000	03:27
Control Test		0.081	03:28
Air Blank		0.000	03:28
Diagnostics Check		OK	03:28

Data Download Date/Time: 02/07/2019 12:53

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HESS,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: SHRIDER

DATE/TIME OF TEST: 01/27/2019 04:01:36

Diagnostics Check		OK	03:54
Air Blank		0.000	03:55
Control Test		0.081	03:55
Air Blank		0.000	03:56
Subject Sample	# 1	0.201	03:56
Breath Volume		2.843	03:56
Air Blank		0.000	03:57
Air Blank		0.000	03:59
Subject Sample	# 2	0.192	04:00
Breath Volume		2.093	04:00
Air Blank		0.000	04:00
Control Test		0.080	04:01
Air Blank		0.000	04:01
Diagnostics Check		OK	04:01

Data Download Date/Time: 02/07/2019 12:54

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SPENCER,K
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: CUMMINGS

DATE/TIME OF TEST: 01/27/2019 06:22:55

Diagnostics Check		OK	06:20	
Air Blank		0.000	06:20	
Control Test		0.081	06:20	
Air Blank		0.000	06:21	
Subject Sample	# 1		06:21	Subject Test Refused
Breath Volume		0.000	06:21	
Air Blank		0.000	06:22	
Control Test		0.081	06:22	
Air Blank		0.000	06:22	
Diagnostics Check		OK	06:22	

Data Download Date/Time: 02/07/2019 12:54

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,F,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DANZIE,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 09:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:30

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 01/27/2019 09:58:05

Diagnostics Check		OK	09:51
Air Blank		0.000	09:51
Control Test		0.082	09:51
Air Blank		0.000	09:52
Subject Sample	# 1	0.131	09:53
Breath Volume		1.515	09:53
Air Blank		0.000	09:53
Air Blank		0.000	09:55
Subject Sample	# 2	0.125	09:56
Breath Volume		1.468	09:56
Air Blank		0.000	09:57
Control Test		0.081	09:57
Air Blank		0.000	09:57
Diagnostics Check		OK	09:58

Data Download Date/Time: 02/07/2019 12:54

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: SEMEDO

DATE/TIME OF TEST: 01/28/2019 00:22:50

Diagnostics Check		OK	00:16
Air Blank		0.000	00:17
Control Test		0.081	00:17
Air Blank		0.000	00:18
Subject Sample	# 1	0.088	00:18
Breath Volume		2.031	00:18
Air Blank		0.000	00:19
Air Blank		0.000	00:21
Subject Sample	# 2	0.083	00:21
Breath Volume		2.578	00:21
Air Blank		0.000	00:21
Control Test		0.081	00:22
Air Blank		0.000	00:22
Diagnostics Check		OK	00:22

Data Download Date/Time: 02/07/2019 12:54

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 01/28/2019 01:47:56

Diagnostics Check		OK	01:45	
Air Blank		0.000	01:45	
Control Test		0.081	01:45	
Air Blank		0.000	01:46	
Subject Sample	# 1		01:46	Subject Test Refused
Breath Volume		0.000	01:46	
Air Blank		0.000	01:47	
Control Test		0.081	01:47	
Air Blank		0.000	01:47	
Diagnostics Check		OK	01:47	

Data Download Date/Time: 02/07/2019 12:54

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: WEBB

DATE/TIME OF TEST: 01/28/2019 02:34:14

Diagnostics Check		OK	02:28
Air Blank		0.000	02:28
Control Test		0.081	02:28
Air Blank		0.000	02:29
Subject Sample	# 1	0.185	02:29
Breath Volume		3.824	02:29
Air Blank		0.000	02:30
Air Blank		0.000	02:32
Subject Sample	# 2	0.179	02:32
Breath Volume		3.035	02:32
Air Blank		0.000	02:33
Control Test		0.081	02:33
Air Blank		0.000	02:34
Diagnostics Check		OK	02:34

Data Download Date/Time: 02/07/2019 12:54

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: FORD

DATE/TIME OF TEST: 01/29/2019 01:20:40

Diagnostics Check		OK	01:14
Air Blank		0.000	01:14
Control Test		0.080	01:15
Air Blank		0.000	01:15
Subject Sample	# 1	0.029	01:16
Breath Volume		2.210	01:16
Air Blank		0.000	01:16
Air Blank		0.000	01:18
Subject Sample	# 2	0.027	01:19
Breath Volume		2.027	01:19
Air Blank		0.000	01:19
Control Test		0.081	01:20
Air Blank		0.000	01:20
Diagnostics Check		OK	01:20

Data Download Date/Time: 02/07/2019 12:54

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: MILEA

DATE/TIME OF TEST: 01/29/2019 03:04:59

Diagnostics Check		OK	02:58
Air Blank		0.000	02:59
Control Test		0.081	02:59
Air Blank		0.000	03:00
Subject Sample	# 1	0.035	03:00
Breath Volume		3.652	03:00
Air Blank		0.000	03:01
Air Blank		0.000	03:03
Subject Sample	# 2	0.037	03:03
Breath Volume		3.878	03:03
Air Blank		0.000	03:04
Control Test		0.081	03:04
Air Blank		0.000	03:04
Diagnostics Check		OK	03:04

Data Download Date/Time: 02/07/2019 12:54

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNS,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: FILLINGIM

DATE/TIME OF TEST: 01/29/2019 04:39:54

Diagnostics Check		OK	04:33
Air Blank		0.000	04:33
Control Test		0.081	04:33
Air Blank		0.000	04:34
Subject Sample	# 1	0.236	04:35
Breath Volume		2.535	04:35
Air Blank		0.000	04:35
Air Blank		0.000	04:37
Subject Sample	# 2	0.224	04:38
Breath Volume		1.714	04:38
Air Blank		0.000	04:38
Control Test		0.081	04:39
Air Blank		0.000	04:39
Diagnostics Check		OK	04:39

Data Download Date/Time: 02/07/2019 12:54

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: WOODLEY

DATE/TIME OF TEST: 01/30/2019 03:48:02

Diagnostics Check		OK	03:41
Air Blank		0.000	03:42
Control Test		0.082	03:42
Air Blank		0.000	03:43
Subject Sample	# 1	0.168	03:43
Breath Volume		3.718	03:43
Air Blank		0.000	03:44
Air Blank		0.000	03:46
Subject Sample	# 2	0.167	03:46
Breath Volume		3.597	03:46
Air Blank		0.000	03:47
Control Test		0.082	03:47
Air Blank		0.000	03:47
Diagnostics Check		OK	03:48

Data Download Date/Time: 02/07/2019 12:54

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: ROZMAN

DATE/TIME OF TEST: 01/31/2019 02:33:15

Diagnostics Check		OK	02:30	
Air Blank		0.000	02:30	
Control Test		0.082	02:31	
Air Blank		0.000	02:31	
Subject Sample	# 1		02:31	Subject Test Refused
Breath Volume		0.000	02:31	
Air Blank		0.000	02:32	
Control Test		0.082	02:32	
Air Blank		0.000	02:33	
Diagnostics Check		OK	02:33	

Data Download Date/Time: 02/07/2019 12:54

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 01/31/2019 04:03:49

Diagnostics Check		OK	04:01	
Air Blank		0.000	04:01	
Control Test		0.082	04:01	
Air Blank		0.000	04:02	
Subject Sample	# 1		04:02	Subject Test Refused
Breath Volume		0.000	04:02	
Air Blank		0.000	04:02	
Control Test		0.081	04:03	
Air Blank		0.000	04:03	
Diagnostics Check		OK	04:03	

Data Download Date/Time: 02/07/2019 12:54

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001284

INSTRUMENT REGISTERED TO: GILCHRIST COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASH,EARL,C
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MCVEIGH,TREVOR
ARRESTING OFFICER AGENCY: Gilchrist County SO
LAST AGENCY INSPECTION: 12/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 19:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: SCHULZ

DATE/TIME OF TEST: 01/02/2019 20:13:46

Diagnostics Check		OK	20:01	
Air Blank		0.000	20:02	
Control Test		0.079	20:02	
Air Blank		0.000	20:02	
Subject Sample	# 1		20:06	No Sample Provided
Breath Volume		0.000	20:06	
Air Blank		0.000	20:06	
Air Blank		0.000	20:08	
Subject Sample	# 2	0.050	20:09	
Breath Volume		3.472	20:09	
Air Blank		0.000	20:09	
Air Blank		0.000	20:11	
Subject Sample	# 3	0.046	20:12	
Breath Volume		3.097	20:12	
Air Blank		0.000	20:12	
Control Test		0.078	20:13	
Air Blank		0.000	20:13	
Diagnostics Check		OK	20:13	

Data Download Date/Time: 02/04/2019 07:45

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001284

INSTRUMENT REGISTERED TO: GILCHRIST COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NETTLES,JONATHAN,K
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: NETTLES,JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 12/24/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 13:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:56

SUBJECT'S LAST NAME: HUGHES

DATE/TIME OF TEST: 01/27/2019 15:35:58

Diagnostics Check		OK	15:29
Air Blank		0.000	15:30
Control Test		0.080	15:30
Air Blank		0.000	15:31
Subject Sample	# 1	0.000	15:31
Breath Volume		3.398	15:31
Air Blank		0.000	15:32
Air Blank		0.000	15:34
Subject Sample	# 2	0.000	15:34
Breath Volume		1.929	15:34
Air Blank		0.000	15:35
Control Test		0.079	15:35
Air Blank		0.000	15:35
Diagnostics Check		OK	15:35

Data Download Date/Time: 02/04/2019 07:45

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

INSTRUMENT REGISTERED TO: BAKER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANDTE,ANDREA,L
BREATH TEST OPERATOR AGENCY: Baker County SO
ARRESTING OFFICER: HIGGINBOTHAM,JEFFERY
ARRESTING OFFICER AGENCY: Baker County SO
LAST AGENCY INSPECTION: 12/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: WEARY JR

DATE/TIME OF TEST: 01/07/2019 22:12:33

Diagnostics Check		OK	22:06
Air Blank		0.000	22:06
Control Test		0.083	22:07
Air Blank		0.000	22:07
Subject Sample	# 1	0.000	22:08
Breath Volume		5.351	22:08
Air Blank		0.000	22:08
Air Blank		0.000	22:10
Subject Sample	# 2	0.000	22:11
Breath Volume		3.093	22:11
Air Blank		0.000	22:11
Control Test		0.083	22:11
Air Blank		0.000	22:12
Diagnostics Check		OK	22:12

Data Download Date/Time: 01/31/2019 23:20

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

INSTRUMENT REGISTERED TO: BAKER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANDTE,ANDREA,L
BREATH TEST OPERATOR AGENCY: Baker County SO
ARRESTING OFFICER: BLOXHAM,T
ARRESTING OFFICER AGENCY: Baker County SO
LAST AGENCY INSPECTION: 12/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: BLUE

DATE/TIME OF TEST: 01/13/2019 03:33:32

Diagnostics Check		OK	03:25
Air Blank		0.000	03:25
Control Test		0.082	03:26
Air Blank		0.000	03:26
Subject Sample	# 1	0.186	03:28
Breath Volume		1.777	03:28
Air Blank		0.000	03:29
Air Blank		0.000	03:31
Subject Sample	# 2	0.183	03:31
Breath Volume		1.238	03:31
Air Blank		0.000	03:32
Control Test		0.081	03:32
Air Blank		0.000	03:33
Diagnostics Check		OK	03:33

Data Download Date/Time: 01/31/2019 23:20

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

INSTRUMENT REGISTERED TO: BAKER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILLIARD,KARI,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HILLIARD,KARI
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 01/13/2019 05:53:23

Diagnostics Check		OK	05:47
Air Blank		0.000	05:47
Control Test		0.083	05:47
Air Blank		0.000	05:48
Subject Sample	# 1	0.000	05:49
Breath Volume		1.671	05:49
Air Blank		0.000	05:49
Air Blank		0.000	05:51
Subject Sample	# 2	0.000	05:52
Breath Volume		1.652	05:52
Air Blank		0.000	05:52
Control Test		0.081	05:52
Air Blank		0.000	05:53
Diagnostics Check		OK	05:53

Data Download Date/Time: 01/31/2019 23:20

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

INSTRUMENT REGISTERED TO: BAKER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILLIARD,KARI,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HILLIARD,KARI
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 14:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: JEWETT

DATE/TIME OF TEST: 01/25/2019 16:02:32

Diagnostics Check		OK	15:56
Air Blank		0.000	15:56
Control Test		0.084	15:57
Air Blank		0.000	15:57
Subject Sample	# 1	0.000	15:58
Breath Volume		3.496	15:58
Air Blank		0.000	15:58
Air Blank		0.000	16:00
Subject Sample	# 2	0.000	16:01
Breath Volume		3.589	16:01
Air Blank		0.000	16:01
Control Test		0.081	16:01
Air Blank		0.000	16:02
Diagnostics Check		OK	16:02

Data Download Date/Time: 01/31/2019 23:20

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VISMAN,CLAYTON,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: VISMAN,CLAYTON
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805701
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: BYRD

DATE/TIME OF TEST: 01/01/2019 03:38:01

Diagnostics Check		OK	03:32	
Air Blank		0.000	03:33	
Control Test		0.077	03:33	
Air Blank		0.000	03:34	
Subject Sample	# 1		03:36	Subject Test Refused
Breath Volume		0.000	03:36	
Air Blank		0.000	03:37	
Control Test		0.077	03:37	
Air Blank		0.000	03:37	
Diagnostics Check		OK	03:38	

Data Download Date/Time: 01/18/2019 08:30

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HULLENDER,ROBERT,S
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: HULLENDER,ROBERT
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 12/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805701
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: STANZE

DATE/TIME OF TEST: 01/12/2019 23:16:58

Diagnostics Check		OK	23:10
Air Blank		0.000	23:10
Control Test		0.079	23:11
Air Blank		0.000	23:11
Subject Sample	# 1	0.227	23:12
Breath Volume		2.355	23:12
Air Blank		0.000	23:13
Air Blank		0.000	23:14
Subject Sample	# 2	0.222	23:15
Breath Volume		1.554	23:15
Air Blank		0.000	23:16
Control Test		0.078	23:16
Air Blank		0.000	23:16
Diagnostics Check		OK	23:16

Data Download Date/Time: 01/18/2019 08:30

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILLIARD,KARI,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HILLIARD,KARI
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 04:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: DOBBS

DATE/TIME OF TEST: 01/17/2019 06:02:33

Diagnostics Check		OK	05:58	
Air Blank		0.000	05:58	
Control Test		0.080	05:58	
Air Blank		0.000	05:59	
Subject Sample	# 1		06:00	Subject Test Refused
Breath Volume		0.085	06:00	
Air Blank		0.000	06:01	
Control Test		0.080	06:01	
Air Blank		0.000	06:02	
Diagnostics Check		OK	06:02	

Data Download Date/Time: 01/18/2019 08:30

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOFMANN,TRAVIS,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HOFMANN,TRAVIS
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: REAGIN

DATE/TIME OF TEST: 01/18/2019 00:54:00

Diagnostics Check		OK	00:50	
Air Blank		0.000	00:51	
Control Test		0.079	00:51	
Air Blank		0.000	00:52	
Subject Sample	# 1		00:52	Subject Test Refused
Breath Volume		0.000	00:52	
Air Blank		0.000	00:53	
Control Test		0.079	00:53	
Air Blank		0.000	00:53	
Diagnostics Check		OK	00:54	

Data Download Date/Time: 01/18/2019 08:30

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,TERRELL,G
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: HANSON,JUSTIN
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 01/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: CARTWRIGHT

DATE/TIME OF TEST: 01/19/2019 00:42:44

Diagnostics Check		OK	00:40	
Air Blank		0.000	00:41	
Control Test		0.080	00:41	
Air Blank		0.000	00:42	
Subject Sample	# 1		00:42	Sequence Aborted
Breath Volume		0.000	00:42	
Air Blank		0.000	00:42	

Data Download Date/Time: 02/10/2019 11:50

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,TERRELL,G
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: HANSON,JUSTIN
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 01/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: CARTWRIGHT

DATE/TIME OF TEST: 01/19/2019 01:11:08

Diagnostics Check		OK	01:04
Air Blank		0.000	01:04
Control Test		0.078	01:04
Air Blank		0.000	01:05
Subject Sample	# 1	0.152	01:06
Breath Volume		1.976	01:06
Air Blank		0.000	01:07
Air Blank		0.000	01:09
Subject Sample	# 2	0.157	01:09
Breath Volume		2.519	01:09
Air Blank		0.000	01:10
Control Test		0.078	01:10
Air Blank		0.000	01:11
Diagnostics Check		OK	01:11

Data Download Date/Time: 02/10/2019 11:50

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HULLENDER,ROBERT,S
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: HULLENDER,ROBERT
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 01/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: SHAFFER

DATE/TIME OF TEST: 01/31/2019 02:44:59

Diagnostics Check		OK	02:38
Air Blank		0.000	02:38
Control Test		0.079	02:38
Air Blank		0.000	02:39
Subject Sample	# 1	0.132	02:40
Breath Volume		3.867	02:40
Air Blank		0.000	02:41
Air Blank		0.000	02:42
Subject Sample	# 2	0.133	02:43
Breath Volume		4.164	02:43
Air Blank		0.000	02:44
Control Test		0.080	02:44
Air Blank		0.000	02:44
Diagnostics Check		OK	02:44

Data Download Date/Time: 02/10/2019 11:50

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KONKEL,SCOTT,M
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: MOSLEY,KIM
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 12/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: PUCKETT

DATE/TIME OF TEST: 01/11/2019 23:01:40

Diagnostics Check		OK	22:56	
Air Blank		0.000	22:56	
Control Test		0.079	22:56	
Air Blank		0.000	22:57	
Subject Sample	# 1	0.016	22:57	
Breath Volume		3.835	22:57	
Air Blank		0.000	22:58	
Air Blank		0.000	23:00	
Subject Sample	# 2	0.014	23:00	
Breath Volume		3.765	23:00	
Air Blank			23:01	RFI Detect
Air Blank		0.000	23:01	

Data Download Date/Time: 01/18/2019 08:32

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUMNER,DALTON,R
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: KROL,FRANKIE
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 12/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: VAN OYAN

DATE/TIME OF TEST: 01/11/2019 23:40:50

Diagnostics Check		OK	23:32	
Air Blank		0.000	23:32	
Control Test		0.079	23:33	
Air Blank		0.000	23:33	
Subject Sample	# 1		23:36	No Sample Provided
Breath Volume		0.000	23:36	
Air Blank		0.000	23:37	
Air Blank		0.000	23:39	
Subject Sample	# 2		23:39	Subject Test Refused
Breath Volume		0.000	23:39	
Air Blank		0.000	23:39	
Control Test		0.079	23:40	
Air Blank		0.000	23:40	
Diagnostics Check		OK	23:40	

Data Download Date/Time: 01/18/2019 08:32

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COFFELT,DYLAN,H
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: COFFELT,DYLAN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 20:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: DONAHOE

DATE/TIME OF TEST: 01/21/2019 21:38:44

Diagnostics Check		OK	21:35	
Air Blank		0.000	21:36	
Control Test		0.080	21:36	
Air Blank		0.000	21:36	
Subject Sample	# 1		21:37	Subject Test Refused
Breath Volume		0.000	21:37	
Air Blank		0.000	21:37	
Control Test		0.080	21:38	
Air Blank		0.000	21:38	
Diagnostics Check		OK	21:38	

Data Download Date/Time: 02/28/2019 17:30

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEINER,ZACHARY,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: STEINER,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 12:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: BROKAW

DATE/TIME OF TEST: 01/22/2019 01:35:04

Diagnostics Check		OK	01:32	
Air Blank		0.000	01:32	
Control Test		0.080	01:32	
Air Blank		0.000	01:33	
Subject Sample	# 1		01:33	Subject Test Refused
Breath Volume		0.000	01:33	
Air Blank		0.000	01:34	
Control Test		0.080	01:34	
Air Blank		0.000	01:34	
Diagnostics Check		OK	01:35	

Data Download Date/Time: 02/28/2019 17:30

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREIRA,JOSE,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: PEREIRA,JOSE
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 18:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:18

SUBJECT'S LAST NAME: CHERNAK

DATE/TIME OF TEST: 01/22/2019 19:45:49

Diagnostics Check		OK	19:39
Air Blank		0.000	19:39
Control Test		0.080	19:39
Air Blank		0.000	19:40
Subject Sample	# 1	0.145	19:41
Breath Volume		2.039	19:41
Air Blank		0.000	19:41
Air Blank		0.000	19:43
Subject Sample	# 2	0.144	19:44
Breath Volume		2.445	19:44
Air Blank		0.000	19:44
Control Test		0.079	19:45
Air Blank		0.000	19:45
Diagnostics Check		OK	19:45

Data Download Date/Time: 02/28/2019 17:30

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,BENJAMIN,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WARD,BENJAMIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 07:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:28

SUBJECT'S LAST NAME: STEVENS

DATE/TIME OF TEST: 01/23/2019 08:24:57

Diagnostics Check		OK	08:22	
Air Blank		0.000	08:22	
Control Test		0.080	08:22	
Air Blank		0.000	08:23	
Subject Sample	# 1		08:23	Subject Test Refused
Breath Volume		0.000	08:23	
Air Blank		0.000	08:24	
Control Test		0.080	08:24	
Air Blank		0.000	08:24	
Diagnostics Check		OK	08:24	

Data Download Date/Time: 02/28/2019 17:30

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JOHN,F
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: BOYD,JACOB
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: CODY

DATE/TIME OF TEST: 01/24/2019 01:01:50

Diagnostics Check		OK	00:46	
Air Blank		0.000	00:46	
Control Test		0.079	00:47	
Air Blank		0.000	00:47	
Subject Sample	# 1	0.212	00:49	
Breath Volume		1.679	00:49	
Air Blank		0.000	00:49	
Air Blank		0.000	00:51	
Subject Sample	# 2		00:54	No Sample Provided
Breath Volume		0.000	00:54	
Air Blank		0.000	00:55	
Air Blank		0.000	00:57	
Subject Sample	# 3		01:00	No Sample Provided
Breath Volume		0.000	01:00	
Air Blank		0.000	01:00	
Control Test		0.079	01:01	
Air Blank		0.000	01:01	
Diagnostics Check		OK	01:01	

Data Download Date/Time: 02/28/2019 17:30

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREIRA,JOSE,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: PEREIRA,JOSE
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 09:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:03

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 01/25/2019 10:29:29

Diagnostics Check		OK	10:23
Air Blank		0.000	10:23
Control Test		0.081	10:24
Air Blank		0.000	10:24
Subject Sample	# 1	0.000	10:25
Breath Volume		1.441	10:25
Air Blank		0.000	10:25
Air Blank		0.000	10:27
Subject Sample	# 2	0.000	10:28
Breath Volume		2.015	10:28
Air Blank		0.000	10:28
Control Test		0.081	10:28
Air Blank		0.000	10:29
Diagnostics Check		OK	10:29

Data Download Date/Time: 02/28/2019 17:30

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOKE,ANTHONY,S
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BAKER,DAVID
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 13:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:28

SUBJECT'S LAST NAME: COOKE

DATE/TIME OF TEST: 01/25/2019 14:59:09

Diagnostics Check		OK	14:56	
Air Blank		0.000	14:56	
Control Test		0.080	14:56	
Air Blank		0.000	14:57	
Subject Sample	# 1		14:57	Interferent Detect
Breath Volume		1.605	14:57	
Air Blank		0.000	14:58	
Control Test		0.081	14:58	
Air Blank		0.000	14:59	
Diagnostics Check		OK	14:59	

Data Download Date/Time: 02/28/2019 17:31

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOKE,ANTHONY,S
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BAKER,DAVID
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 13:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:28

SUBJECT'S LAST NAME: COOKE

DATE/TIME OF TEST: 01/25/2019 15:06:10

Diagnostics Check		OK	15:02	
Air Blank		0.000	15:03	
Control Test		0.080	15:03	
Air Blank		0.000	15:04	
Subject Sample	# 1		15:04	Interferent Detect
Breath Volume		2.757	15:04	
Air Blank		0.000	15:05	
Control Test		0.081	15:05	
Air Blank		0.000	15:06	
Diagnostics Check		OK	15:06	

Data Download Date/Time: 02/28/2019 17:31

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAPMAN,MAVERICK,E
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CHAPMAN,MAVERICK
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 19:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: ROHER

DATE/TIME OF TEST: 01/25/2019 20:25:29

Diagnostics Check		OK	20:19
Air Blank		0.000	20:19
Control Test		0.080	20:20
Air Blank		0.000	20:20
Subject Sample	# 1	0.227	20:21
Breath Volume		1.261	20:21
Air Blank		0.000	20:21
Air Blank		0.000	20:23
Subject Sample	# 2	0.221	20:23
Breath Volume		1.695	20:23
Air Blank		0.000	20:24
Control Test		0.080	20:24
Air Blank		0.000	20:25
Diagnostics Check		OK	20:25

Data Download Date/Time: 02/28/2019 17:31

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONDURANT,JASON,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: BONDURANT,JASON
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 14:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:33

SUBJECT'S LAST NAME: HUNTER

DATE/TIME OF TEST: 01/26/2019 15:59:53

Diagnostics Check		OK	15:53
Air Blank		0.000	15:54
Control Test		0.080	15:54
Air Blank		0.000	15:55
Subject Sample	# 1	0.151	15:55
Breath Volume		1.238	15:55
Air Blank		0.000	15:56
Air Blank		0.000	15:57
Subject Sample	# 2	0.162	15:58
Breath Volume		2.847	15:58
Air Blank		0.000	15:58
Control Test		0.079	15:59
Air Blank		0.000	15:59
Diagnostics Check		OK	15:59

Data Download Date/Time: 02/28/2019 17:31

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWTON,CARSON,H
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: GRAY,ANN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 04:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:54

SUBJECT'S LAST NAME: BLOUNT

DATE/TIME OF TEST: 01/27/2019 05:27:29

Diagnostics Check		OK	05:19	
Air Blank		0.000	05:19	
Control Test		0.080	05:19	
Air Blank		0.000	05:20	
Subject Sample	# 1		05:23	No Sample Provided
Breath Volume		0.000	05:23	
Air Blank		0.000	05:23	
Air Blank		0.000	05:25	
Subject Sample	# 2		05:26	Subject Test Refused
Breath Volume		0.000	05:26	
Air Blank		0.000	05:26	
Control Test		0.080	05:26	
Air Blank		0.000	05:27	
Diagnostics Check		OK	05:27	

Data Download Date/Time: 02/28/2019 17:31

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNGREN,DENNIS,A
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: YOUNGREN,DENNIS
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 01/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 22:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: BOURG

DATE/TIME OF TEST: 01/29/2019 00:38:37

Diagnostics Check		OK	00:32
Air Blank		0.000	00:32
Control Test		0.080	00:33
Air Blank		0.000	00:33
Subject Sample	# 1	0.000	00:34
Breath Volume		2.554	00:34
Air Blank		0.000	00:34
Air Blank		0.000	00:36
Subject Sample	# 2	0.000	00:37
Breath Volume		2.574	00:37
Air Blank		0.000	00:37
Control Test		0.080	00:38
Air Blank		0.000	00:38
Diagnostics Check		OK	00:38

Data Download Date/Time: 02/28/2019 17:31

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEINER,ZACHARY,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: STEINER,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 16:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:10

SUBJECT'S LAST NAME: KANE

DATE/TIME OF TEST: 01/29/2019 17:37:38

Diagnostics Check		OK	17:30
Air Blank		0.000	17:31
Control Test		0.081	17:31
Air Blank		0.000	17:31
Subject Sample	# 1	0.135	17:33
Breath Volume		1.355	17:33
Air Blank		0.000	17:33
Air Blank		0.000	17:35
Subject Sample	# 2	0.147	17:36
Breath Volume		2.750	17:36
Air Blank		0.000	17:36
Control Test		0.080	17:37
Air Blank		0.000	17:37
Diagnostics Check		OK	17:37

Data Download Date/Time: 02/28/2019 17:31

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,CALEB,Z
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: DELANEY,TYLER
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: MCQUAIGE

DATE/TIME OF TEST: 01/29/2019 23:32:02

Diagnostics Check		OK	23:25
Air Blank		0.000	23:25
Control Test		0.081	23:26
Air Blank		0.000	23:26
Subject Sample	# 1	0.000	23:27
Breath Volume		2.378	23:27
Air Blank		0.000	23:27
Air Blank		0.000	23:29
Subject Sample	# 2	0.000	23:30
Breath Volume		1.558	23:30
Air Blank		0.000	23:31
Control Test		0.081	23:31
Air Blank		0.000	23:31
Diagnostics Check		OK	23:32

Data Download Date/Time: 02/28/2019 17:31

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,ERIC,D
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: BROWN,ERIC
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 12/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG722402
DRY GAS EXPIRY DATE: 08/12/2019
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: SURRATT

DATE/TIME OF TEST: 01/01/2019 03:50:10

Diagnostics Check		OK	03:44
Air Blank		0.000	03:44
Control Test		0.079	03:45
Air Blank		0.000	03:45
Subject Sample	# 1	0.122	03:45
Breath Volume		1.191	03:45
Air Blank		0.000	03:46
Air Blank		0.000	03:48
Subject Sample	# 2	0.118	03:48
Breath Volume		1.523	03:48
Air Blank		0.000	03:49
Control Test		0.080	03:49
Air Blank		0.000	03:50
Diagnostics Check		OK	03:50

Data Download Date/Time: 01/31/2019 12:00

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROSIUS,STANLEY,W
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: ROBERTS,B
ARRESTING OFFICER AGENCY: Springfield PD
LAST AGENCY INSPECTION: 12/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG722402
DRY GAS EXPIRY DATE: 08/12/2019
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: KEEL

DATE/TIME OF TEST: 01/01/2019 04:31:42

Diagnostics Check		OK	04:24
Air Blank		0.000	04:24
Control Test		0.079	04:25
Air Blank		0.000	04:25
Subject Sample	# 1	0.219	04:26
Breath Volume		1.527	04:26
Air Blank		0.000	04:26
Air Blank		0.000	04:28
Subject Sample	# 2	0.225	04:29
Breath Volume		3.039	04:29
Air Blank		0.000	04:30
Control Test		0.076	04:31
Air Blank		0.000	04:31
Diagnostics Check		OK	04:31

Data Download Date/Time: 01/31/2019 12:00

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINCOLN,TREY,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LINCOLN,TREY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG722402
DRY GAS EXPIRY DATE: 08/12/2019
VIOLATION CODE: DUI
ARREST TIME: 14:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:38

SUBJECT'S LAST NAME: MAYNOR

DATE/TIME OF TEST: 01/03/2019 15:02:25

Diagnostics Check	OK	15:01	
Air Blank	0.000	15:01	
Control Test	0.071	15:01	RFI Detect
Air Blank	0.000	15:02	

Data Download Date/Time: 01/31/2019 12:00

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINCOLN,TREY,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LINCOLN,TREY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG722402
DRY GAS EXPIRY DATE: 08/12/2019
VIOLATION CODE: DUI
ARREST TIME: 14:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:38

SUBJECT'S LAST NAME: MAYNOR

DATE/TIME OF TEST: 01/03/2019 15:12:45

Diagnostics Check		OK	15:06
Air Blank		0.000	15:06
Control Test		0.079	15:06
Air Blank		0.000	15:07
Subject Sample	# 1	0.000	15:08
Breath Volume		1.832	15:08
Air Blank		0.000	15:08
Air Blank		0.000	15:10
Subject Sample	# 2	0.000	15:11
Breath Volume		2.515	15:11
Air Blank		0.000	15:11
Control Test		0.079	15:12
Air Blank		0.000	15:12
Diagnostics Check		OK	15:12

Data Download Date/Time: 01/31/2019 12:00

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCE,BRAD,
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: ROBERTS,J
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 12/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG722402
DRY GAS EXPIRY DATE: 08/12/2019
VIOLATION CODE: DUI
ARREST TIME: 14:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:40

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 01/04/2019 15:12:28

Diagnostics Check		OK	15:00	
Air Blank		0.000	15:00	
Control Test		0.080	15:01	
Air Blank		0.000	15:01	
Subject Sample	# 1		15:04	No Sample Provided
Breath Volume		0.000	15:04	
Air Blank		0.000	15:05	
Air Blank		0.000	15:07	
Subject Sample	# 2	0.248	15:07	
Breath Volume		1.460	15:07	
Air Blank		0.000	15:08	
Air Blank		0.000	15:10	
Subject Sample	# 3	0.243	15:10	
Breath Volume		1.781	15:10	
Air Blank		0.000	15:11	
Control Test		0.079	15:11	
Air Blank		0.000	15:12	
Diagnostics Check		OK	15:12	

Data Download Date/Time: 01/31/2019 12:00

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOHRER,MARK,
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: ROSSOMONDO,MIKE
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 12/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG722402
DRY GAS EXPIRY DATE: 08/12/2019
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 01/05/2019 04:06:58

Diagnostics Check		OK	04:00
Air Blank		0.000	04:01
Control Test		0.080	04:01
Air Blank		0.000	04:02
Subject Sample	# 1	0.118	04:02
Breath Volume		1.957	04:02
Air Blank		0.000	04:03
Air Blank		0.000	04:05
Subject Sample	# 2	0.104	04:05
Breath Volume		1.378	04:05
Air Blank		0.000	04:06
Control Test		0.079	04:06
Air Blank		0.000	04:06
Diagnostics Check		OK	04:06

Data Download Date/Time: 01/31/2019 12:00

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,TIMOTHY,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LEWIS,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG722402
DRY GAS EXPIRY DATE: 08/12/2019
VIOLATION CODE: DUI
ARREST TIME: 15:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:32

SUBJECT'S LAST NAME: ELIAS

DATE/TIME OF TEST: 01/10/2019 15:58:22

Diagnostics Check		OK	15:52	
Air Blank		0.000	15:53	
Control Test		0.080	15:53	
Air Blank		0.000	15:54	
Subject Sample	# 1	0.000	15:54	
Breath Volume		2.050	15:54	
Air Blank		0.000	15:54	
Air Blank		0.000	15:56	
Subject Sample	# 2	0.000	15:57	
Breath Volume		1.875	15:57	
Air Blank		0.000	15:57	
			15:57	RFI Detect
Air Blank		0.000	15:58	

Data Download Date/Time: 01/31/2019 12:00

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,RYAN,D
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: ADAMS,HOLTON
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 12/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG722402
DRY GAS EXPIRY DATE: 08/12/2019
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 01/17/2019 22:58:50

Diagnostics Check		OK	22:52
Air Blank		0.000	22:52
Control Test		0.081	22:52
Air Blank		0.000	22:53
Subject Sample	# 1	0.000	22:54
Breath Volume		1.722	22:54
Air Blank		0.000	22:54
Air Blank		0.000	22:56
Subject Sample	# 2	0.000	22:57
Breath Volume		1.457	22:57
Air Blank		0.000	22:57
Control Test		0.080	22:58
Air Blank		0.000	22:58
Diagnostics Check		OK	22:58

Data Download Date/Time: 01/31/2019 12:00

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRELL,CHAD,
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: HARRELL,CHAD
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 12/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG722402
DRY GAS EXPIRY DATE: 08/12/2019
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: DUARTE

DATE/TIME OF TEST: 01/19/2019 02:32:35

Diagnostics Check		OK	02:24
Air Blank		0.000	02:24
Control Test		0.080	02:24
Air Blank		0.000	02:25
Subject Sample	# 1	0.257	02:28
Breath Volume		3.152	02:28
Air Blank		0.000	02:28
Air Blank		0.000	02:30
Subject Sample	# 2	0.254	02:31
Breath Volume		2.500	02:31
Air Blank		0.000	02:31
Control Test		0.079	02:31
Air Blank		0.000	02:32
Diagnostics Check		OK	02:32

Data Download Date/Time: 01/31/2019 12:00

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWMAN,JOANNA,N
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BOWMAN,JOANNA
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG722402
DRY GAS EXPIRY DATE: 08/12/2019
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: CODY

DATE/TIME OF TEST: 01/20/2019 03:11:57

Diagnostics Check		OK	03:05
Air Blank		0.000	03:05
Control Test		0.080	03:06
Air Blank		0.000	03:06
Subject Sample	# 1	0.097	03:07
Breath Volume		1.285	03:07
Air Blank		0.000	03:07
Air Blank		0.000	03:09
Subject Sample	# 2	0.100	03:10
Breath Volume		1.527	03:10
Air Blank		0.000	03:11
Control Test		0.080	03:11
Air Blank		0.000	03:11
Diagnostics Check		OK	03:11

Data Download Date/Time: 01/31/2019 12:00

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WITHROW,TRACY,L
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: PALO,THERESA
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 12/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG722402
DRY GAS EXPIRY DATE: 08/12/2019
VIOLATION CODE: DUI
ARREST TIME: 05:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:05

SUBJECT'S LAST NAME: SHEA

DATE/TIME OF TEST: 01/23/2019 06:31:09

Diagnostics Check		OK	06:25
Air Blank		0.000	06:25
Control Test		0.080	06:25
Air Blank		0.000	06:26
Subject Sample	# 1	0.000	06:26
Breath Volume		3.960	06:26
Air Blank		0.000	06:27
Air Blank		0.000	06:29
Subject Sample	# 2	0.000	06:29
Breath Volume		3.371	06:29
Air Blank		0.000	06:30
Control Test		0.079	06:30
Air Blank		0.000	06:31
Diagnostics Check		OK	06:31

Data Download Date/Time: 01/31/2019 12:00

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUBEL,NEIL,F
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RUBEL,NEIL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG722402
DRY GAS EXPIRY DATE: 08/12/2019
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: MIMS

DATE/TIME OF TEST: 01/25/2019 01:42:46

Diagnostics Check		OK	01:39	
Air Blank		0.000	01:39	
Control Test		0.077	01:40	
Air Blank		0.000	01:40	
Subject Sample	# 1		01:41	Subject Test Refused
Breath Volume		0.000	01:41	
Air Blank		0.000	01:41	
Control Test		0.076	01:42	
Air Blank		0.000	01:42	
Diagnostics Check		OK	01:42	

Data Download Date/Time: 01/31/2019 12:00

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,TIMOTHY,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LEWIS,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG722402
DRY GAS EXPIRY DATE: 08/12/2019
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 01/25/2019 03:32:19

Diagnostics Check		OK	03:26
Air Blank		0.000	03:26
Control Test		0.081	03:26
Air Blank		0.000	03:27
Subject Sample	# 1	0.138	03:27
Breath Volume		4.281	03:27
Air Blank		0.000	03:28
Air Blank		0.000	03:30
Subject Sample	# 2	0.138	03:30
Breath Volume		3.882	03:30
Air Blank		0.000	03:31
Control Test		0.080	03:31
Air Blank		0.000	03:32
Diagnostics Check		OK	03:32

Data Download Date/Time: 01/31/2019 12:00

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUBEL,NEIL,F
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RUBEL,NEIL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG722402
DRY GAS EXPIRY DATE: 08/12/2019
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: HARTZELL

DATE/TIME OF TEST: 01/27/2019 01:52:15

Diagnostics Check		OK	01:45
Air Blank		0.000	01:46
Control Test		0.081	01:46
Air Blank		0.000	01:47
Subject Sample	# 1	0.229	01:47
Breath Volume		2.343	01:47
Air Blank		0.000	01:48
Air Blank		0.000	01:50
Subject Sample	# 2	0.229	01:50
Breath Volume		2.140	01:50
Air Blank		0.000	01:51
Control Test		0.080	01:51
Air Blank		0.000	01:52
Diagnostics Check		OK	01:52

Data Download Date/Time: 01/31/2019 12:00

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUBEL,NEIL,F
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RUBEL,NEIL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG722402
DRY GAS EXPIRY DATE: 08/12/2019
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 01/29/2019 03:36:29

Diagnostics Check		OK	03:33	
Air Blank		0.000	03:33	
Control Test		0.080	03:34	
Air Blank		0.000	03:34	
Subject Sample	# 1		03:35	Subject Test Refused
Breath Volume		0.000	03:35	
Air Blank		0.000	03:35	
Control Test		0.080	03:35	
Air Blank		0.000	03:36	
Diagnostics Check		OK	03:36	

Data Download Date/Time: 01/31/2019 12:00

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUTTON,KERRY,D
BREATH TEST OPERATOR AGENCY: Lynn Haven PD
ARRESTING OFFICER: DUTTON,KERRY
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 01/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG722402
DRY GAS EXPIRY DATE: 08/12/2019
VIOLATION CODE: DUI
ARREST TIME: 12:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:35

SUBJECT'S LAST NAME: PENYAK

DATE/TIME OF TEST: 01/31/2019 13:03:10

Diagnostics Check		OK	12:57	
Air Blank		0.000	12:58	
Control Test		0.079	12:58	
Air Blank		0.000	12:58	
Subject Sample	# 1		13:01	Subject Test Refused
Breath Volume		0.113	13:01	
Air Blank		0.000	13:02	
Control Test		0.082	13:02	
Air Blank		0.000	13:03	
Diagnostics Check		OK	13:03	

Data Download Date/Time: 02/27/2019 11:43

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORDHAM,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: CARRICO,KELLY
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 12/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 869968
DRY GAS EXPIRY DATE: 12/06/2020
VIOLATION CODE: DUI
ARREST TIME: 05:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:05

SUBJECT'S LAST NAME: LOGSDON

DATE/TIME OF TEST: 01/09/2019 18:05:43

Diagnostics Check	OK	18:04	
Air Blank	0.000	18:04	
Control Test	0.081	18:04	
Air Blank		18:05	Sequence Aborted
Air Blank	0.000	18:05	

Data Download Date/Time: 02/04/2019 23:45

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORDHAM,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: CARRICO,KELLY
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 12/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 869968
DRY GAS EXPIRY DATE: 12/06/2020
VIOLATION CODE: DUI
ARREST TIME: 17:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:06

SUBJECT'S LAST NAME: LOGSDON

DATE/TIME OF TEST: 01/09/2019 18:32:48

Diagnostics Check		OK	18:26
Air Blank		0.000	18:27
Control Test		0.081	18:27
Air Blank		0.000	18:27
Subject Sample	# 1	0.000	18:28
Breath Volume		2.937	18:28
Air Blank		0.000	18:28
Air Blank		0.000	18:30
Subject Sample	# 2	0.000	18:31
Breath Volume		3.324	18:31
Air Blank		0.000	18:31
Control Test		0.082	18:32
Air Blank		0.000	18:32
Diagnostics Check		OK	18:32

Data Download Date/Time: 02/04/2019 23:45

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCINTOSH,ZANE,R
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: MCINTOSH,ZANE
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 12/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 171502
DRY GAS EXPIRY DATE: 12/06/2020
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 01/09/2019 23:14:50

Diagnostics Check		OK	23:12	
Air Blank		0.000	23:12	
Control Test		0.081	23:12	
Air Blank		0.000	23:13	
Subject Sample	# 1		23:13	Subject Test Refused
Breath Volume		0.000	23:13	
Air Blank		0.000	23:13	
Control Test		0.082	23:14	
Air Blank		0.000	23:14	
Diagnostics Check		OK	23:14	

Data Download Date/Time: 02/04/2019 23:45

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEWBY,JOHN,O
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: RICHARD,MICHAEL
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 01/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 171502
DRY GAS EXPIRY DATE: 12/06/2020
VIOLATION CODE: DUI
ARREST TIME: 18:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: MATTHEWS

DATE/TIME OF TEST: 01/31/2019 19:51:34

Diagnostics Check		OK	19:48	
Air Blank		0.000	19:49	
Control Test		0.082	19:49	
Air Blank		0.000	19:50	
Subject Sample	# 1		19:50	Subject Test Refused
Breath Volume		0.000	19:50	
Air Blank		0.000	19:50	
Control Test		0.082	19:50	
Air Blank		0.000	19:51	
Diagnostics Check		OK	19:51	

Data Download Date/Time: 02/04/2019 23:45

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHULTZ,JEREMY,J
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: SCHULTZ,JEREMY
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 12/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: HAYES

DATE/TIME OF TEST: 01/01/2019 00:35:30

Diagnostics Check		OK	00:24	
Air Blank		0.000	00:24	
Control Test		0.079	00:24	
Air Blank		0.000	00:25	
Subject Sample	# 1		00:28	No Sample Provided
Breath Volume		0.000	00:28	
Air Blank		0.000	00:28	
Air Blank		0.000	00:30	
Subject Sample	# 2		00:34	No Sample Provided
Breath Volume		0.000	00:34	
Air Blank		0.000	00:34	
Control Test		0.080	00:34	
Air Blank		0.000	00:35	
Diagnostics Check		OK	00:35	

Data Download Date/Time: 01/23/2019 13:09

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCORMICK,CHANTEL,B
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: MCCORMICK,CHANTEL
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 12/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 04:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 01/01/2019 05:19:22

Diagnostics Check	OK	05:18	
Air Blank	0.000	05:18	
Control Test	0.070	05:18	Control Outside Tolerance
Air Blank	0.000	05:19	

Data Download Date/Time: 01/23/2019 13:09

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCORMICK,CHANTEL,B
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: MCCORMICK,CHANTEL
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 12/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 04:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 01/01/2019 05:28:01

Diagnostics Check		OK	05:21
Air Blank		0.000	05:21
Control Test		0.079	05:21
Air Blank		0.000	05:22
Subject Sample	# 1	0.171	05:23
Breath Volume		1.828	05:23
Air Blank		0.000	05:24
Air Blank		0.000	05:25
Subject Sample	# 2	0.174	05:26
Breath Volume		1.984	05:26
Air Blank		0.000	05:27
Control Test		0.075	05:27
Air Blank		0.000	05:27
Diagnostics Check		OK	05:28

Data Download Date/Time: 01/23/2019 13:09

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,DAVID,O
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: COOK,CHRISTOPHER
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 12/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 17:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: VAN

DATE/TIME OF TEST: 01/01/2019 19:38:58

Diagnostics Check		OK	19:35	
Air Blank		0.000	19:36	
Control Test		0.079	19:36	
Air Blank		0.000	19:37	
Subject Sample	# 1		19:37	Subject Test Refused
Breath Volume		0.000	19:37	
Air Blank		0.000	19:38	
Control Test		0.079	19:38	
Air Blank		0.000	19:38	
Diagnostics Check		OK	19:38	

Data Download Date/Time: 01/23/2019 13:09

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAYDOS,JONATHAN,A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: MCCOY,THOMAS
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 12/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: COURT ORD
ARREST TIME: 10:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:23

SUBJECT'S LAST NAME: GARRETT

DATE/TIME OF TEST: 01/02/2019 10:53:20

Diagnostics Check		OK	10:46
Air Blank		0.000	10:46
Control Test		0.080	10:47
Air Blank		0.000	10:47
Subject Sample	# 1	0.354	10:48
Breath Volume		1.546	10:48
Air Blank		0.000	10:48
Air Blank		0.000	10:50
Subject Sample	# 2	0.371	10:51
Breath Volume		1.351	10:51
Air Blank		0.000	10:52
Control Test		0.081	10:52
Air Blank		0.000	10:53
Diagnostics Check		OK	10:53

Data Download Date/Time: 01/23/2019 13:09

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUTTO,KYLE,K
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: VENTRY,WILLIS
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 12/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: VIRGIN

DATE/TIME OF TEST: 01/13/2019 21:36:50

Diagnostics Check		OK	21:27	
Air Blank		0.000	21:28	
Control Test		0.079	21:28	
Air Blank		0.000	21:28	
Subject Sample	# 1		21:31	No Sample Provided
Breath Volume		0.000	21:31	
Air Blank		0.000	21:32	
Air Blank		0.000	21:34	
Subject Sample	# 2		21:35	Subject Test Refused
Breath Volume		0.000	21:35	
Air Blank		0.000	21:35	
Control Test		0.080	21:36	
Air Blank		0.000	21:36	
Diagnostics Check		OK	21:36	

Data Download Date/Time: 01/23/2019 13:09

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAY,JODY,L
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: DAY,JODY
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 12/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: OAKS

DATE/TIME OF TEST: 01/14/2019 02:19:29

Diagnostics Check		OK	02:17	
Air Blank		0.000	02:17	
Control Test		0.079	02:17	
Air Blank		0.000	02:18	
Subject Sample	# 1		02:18	RFI Detect
Breath Volume		0.000	02:18	
Air Blank			02:19	RFI Detect
Air Blank		0.000	02:19	

Data Download Date/Time: 01/23/2019 13:09

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAY,JODY,L
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: DAY,JODY
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 12/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: OAKS

DATE/TIME OF TEST: 01/14/2019 02:29:35

Diagnostics Check		OK	02:21	
Air Blank		0.000	02:22	
Control Test		0.079	02:22	
Air Blank		0.000	02:22	
Subject Sample	# 1	0.215	02:23	
Breath Volume		1.804	02:23	
Air Blank		0.000	02:23	
Air Blank		0.000	02:25	
Subject Sample	# 2	0.212	02:27	
Breath Volume		2.273	02:27	
Air Blank		0.000	02:28	
Control Test		0.080	02:28	
Air Blank			02:29	RFI Detect
Air Blank		0.000	02:29	

Data Download Date/Time: 01/23/2019 13:09

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVERY,JOHN,W
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: BRANNON,JUSTIN
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 12/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 01/15/2019 00:17:30

Diagnostics Check		OK	00:10
Air Blank		0.000	00:11
Control Test		0.078	00:11
Air Blank		0.000	00:12
Subject Sample	# 1	0.191	00:12
Breath Volume		3.738	00:12
Air Blank		0.000	00:13
Air Blank		0.000	00:15
Subject Sample	# 2	0.196	00:15
Breath Volume		3.796	00:15
Air Blank		0.000	00:16
Control Test		0.079	00:16
Air Blank		0.000	00:17
Diagnostics Check		OK	00:17

Data Download Date/Time: 01/23/2019 13:09

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVERY,JOHN,W
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: CAREY,ANDREW
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 12/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: BRUCE

DATE/TIME OF TEST: 01/19/2019 02:07:30

Diagnostics Check		OK	02:04	
Air Blank		0.000	02:04	
Control Test		0.079	02:05	
Air Blank		0.000	02:05	
Subject Sample	# 1		02:06	Subject Test Refused
Breath Volume		0.000	02:06	
Air Blank		0.000	02:06	
Control Test		0.080	02:06	
Air Blank		0.000	02:07	
Diagnostics Check		OK	02:07	

Data Download Date/Time: 01/23/2019 13:09

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVERY,JOHN,W
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: ARNOLD,ALEXIS
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 12/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 18:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:33

SUBJECT'S LAST NAME: FREDRICK

DATE/TIME OF TEST: 01/19/2019 20:06:08

Diagnostics Check		OK	19:55
Air Blank		0.000	19:56
Control Test		0.079	19:56
Air Blank		0.000	19:57
Subject Sample	# 1	0.377	19:57
Breath Volume		2.390	19:57
Air Blank		0.000	19:58
Air Blank		0.000	20:00
Subject Sample	# 2	0.351	20:01
Breath Volume		1.726	20:01
Air Blank		0.000	20:01
Air Blank		0.000	20:03
Subject Sample	# 3	0.345	20:04
Breath Volume		2.050	20:04
Air Blank		0.000	20:05
Control Test		0.079	20:05
Air Blank		0.000	20:06
Diagnostics Check		OK	20:06

Data Download Date/Time: 01/23/2019 13:09

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENNETT,DAVID,J
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: ARNOLD,ALEXI
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 12/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: COURT ORD
ARREST TIME: 19:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:10

SUBJECT'S LAST NAME: FREDRICK

DATE/TIME OF TEST: 01/20/2019 11:44:55

Diagnostics Check		OK	11:38
Air Blank		0.000	11:39
Control Test		0.081	11:39
Air Blank		0.000	11:40
Subject Sample	# 1	0.022	11:40
Breath Volume		2.734	11:40
Air Blank		0.000	11:41
Air Blank		0.000	11:42
Subject Sample	# 2	0.018	11:43
Breath Volume		2.609	11:43
Air Blank		0.000	11:44
Control Test		0.079	11:44
Air Blank		0.000	11:44
Diagnostics Check		OK	11:44

Data Download Date/Time: 01/23/2019 13:09

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVERY,JOHN,W
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: COKONOUGHER,ROBERT
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 12/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 20:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: SALMON

DATE/TIME OF TEST: 01/20/2019 22:35:32

Diagnostics Check		OK	22:29
Air Blank		0.000	22:29
Control Test		0.079	22:30
Air Blank		0.000	22:30
Subject Sample	# 1	0.139	22:31
Breath Volume		2.015	22:31
Air Blank		0.000	22:31
Air Blank		0.000	22:33
Subject Sample	# 2	0.151	22:34
Breath Volume		1.507	22:34
Air Blank		0.000	22:34
Control Test		0.080	22:34
Air Blank		0.000	22:35
Diagnostics Check		OK	22:35

Data Download Date/Time: 01/23/2019 13:09

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,DAVID,O
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: MCGRAW,JESSICA
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 01/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 17:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: DUPREE

DATE/TIME OF TEST: 01/24/2019 19:24:04

Diagnostics Check		OK	19:20	
Air Blank		0.000	19:21	
Control Test		0.080	19:21	
Air Blank		0.000	19:22	
Subject Sample	# 1		19:22	Subject Test Refused
Breath Volume		0.000	19:22	
Air Blank		0.000	19:23	
Control Test		0.080	19:23	
Air Blank		0.000	19:23	
Diagnostics Check		OK	19:24	

Data Download Date/Time: 02/05/2019 09:10

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAY,JODY,L
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: DAY,JODY
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 01/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 22:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: MCCALL

DATE/TIME OF TEST: 01/30/2019 23:05:58

Diagnostics Check		OK	22:57
Air Blank		0.000	22:57
Control Test		0.080	22:57
Air Blank		0.000	22:58
Subject Sample	# 1	0.212	23:00
Breath Volume		1.457	23:00
Air Blank		0.000	23:01
Air Blank		0.000	23:02
Subject Sample	# 2	0.210	23:04
Breath Volume		1.457	23:04
Air Blank		0.000	23:05
Control Test		0.079	23:05
Air Blank		0.000	23:05
Diagnostics Check		OK	23:05

Data Download Date/Time: 02/05/2019 09:10

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUNN,CONNOR,S
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: GUNN,CONNOR
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 12/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: LEYBA

DATE/TIME OF TEST: 01/12/2019 00:37:13

Diagnostics Check		OK	00:28	
Air Blank		0.000	00:28	
Control Test		0.081	00:28	
Air Blank		0.000	00:29	
Subject Sample	# 1		00:32	No Sample Provided
Breath Volume		0.000	00:32	
Air Blank		0.000	00:32	
Air Blank		0.000	00:34	
Subject Sample	# 2		00:35	Subject Test Refused
Breath Volume		0.000	00:35	
Air Blank		0.000	00:36	
Control Test		0.081	00:36	
Air Blank		0.000	00:37	
Diagnostics Check		OK	00:37	

Data Download Date/Time: 01/23/2019 13:18

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAYDOS,SAVANNAH,M
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: YOST,CRAIG
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 12/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 13:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: PEREZ-LOPEZ

DATE/TIME OF TEST: 01/12/2019 14:36:34

Diagnostics Check		OK	14:29
Air Blank		0.000	14:30
Control Test		0.080	14:30
Air Blank		0.000	14:31
Subject Sample	# 1	0.300	14:31
Breath Volume		3.039	14:31
Air Blank		0.000	14:32
Air Blank		0.000	14:34
Subject Sample	# 2	0.305	14:34
Breath Volume		1.503	14:34
Air Blank		0.000	14:35
Control Test		0.079	14:35
Air Blank		0.000	14:36
Diagnostics Check		OK	14:36

Data Download Date/Time: 01/23/2019 13:18

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUNN,CONNOR,S
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: MCGHEE,DARRYL
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 12/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 19:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: PURVIS

DATE/TIME OF TEST: 01/13/2019 21:07:08

Diagnostics Check		OK	21:03	
Air Blank		0.000	21:03	
Control Test		0.081	21:04	
Air Blank		0.000	21:04	
Subject Sample	# 1		21:05	Subject Test Refused
Breath Volume		0.000	21:05	
Air Blank		0.000	21:06	
Control Test		0.082	21:06	
Air Blank		0.000	21:07	
Diagnostics Check		OK	21:07	

Data Download Date/Time: 01/23/2019 13:18

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUNN,CONNOR,S
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: OTILUS,LEBEDOPH
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 12/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 19:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: HUSTON

DATE/TIME OF TEST: 01/13/2019 21:32:29

Diagnostics Check		OK	21:27	
Air Blank		0.000	21:28	
Control Test		0.080	21:28	
Air Blank		0.000	21:29	
Subject Sample	# 1		21:31	Subject Test Refused
Breath Volume		0.000	21:31	
Air Blank		0.000	21:31	
Control Test		0.081	21:31	
Air Blank		0.000	21:32	
Diagnostics Check		OK	21:32	

Data Download Date/Time: 01/23/2019 13:18

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUNN,CONNOR,S
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: GUNN,CONNOR
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 12/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: MCGOWAN

DATE/TIME OF TEST: 01/18/2019 01:21:32

Diagnostics Check		OK	01:15
Air Blank		0.000	01:15
Control Test		0.081	01:16
Air Blank		0.000	01:16
Subject Sample	# 1	0.200	01:17
Breath Volume		1.761	01:17
Air Blank		0.000	01:17
Air Blank		0.000	01:19
Subject Sample	# 2	0.202	01:19
Breath Volume		1.820	01:19
Air Blank		0.000	01:20
Control Test		0.081	01:20
Air Blank		0.000	01:21
Diagnostics Check		OK	01:21

Data Download Date/Time: 01/23/2019 13:18

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,B,
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: MILLER,B
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 12/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: PROB/PAROL
ARREST TIME: 11:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:25

SUBJECT'S LAST NAME: LANASA

DATE/TIME OF TEST: 01/19/2019 12:04:56

Diagnostics Check		OK	11:58
Air Blank		0.000	11:59
Control Test		0.080	11:59
Air Blank		0.000	11:59
Subject Sample	# 1	0.000	12:00
Breath Volume		2.406	12:00
Air Blank		0.000	12:00
Air Blank		0.000	12:02
Subject Sample	# 2	0.000	12:03
Breath Volume		2.000	12:03
Air Blank		0.000	12:04
Control Test		0.078	12:04
Air Blank		0.000	12:04
Diagnostics Check		OK	12:04

Data Download Date/Time: 01/23/2019 13:18

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,STEVEN,D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK,STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: CLINE

DATE/TIME OF TEST: 01/05/2019 02:55:17

Diagnostics Check		OK	02:48
Air Blank		0.000	02:48
Control Test		0.080	02:48
Air Blank		0.000	02:49
Subject Sample	# 1	0.175	02:50
Breath Volume		2.656	02:50
Air Blank		0.000	02:51
Air Blank		0.000	02:53
Subject Sample	# 2	0.165	02:53
Breath Volume		1.882	02:53
Air Blank		0.000	02:54
Control Test		0.080	02:54
Air Blank		0.000	02:55
Diagnostics Check		OK	02:55

Data Download Date/Time: 01/08/2019 08:56

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,STEVEN,D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK,STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 01/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: VITELLA

DATE/TIME OF TEST: 01/12/2019 00:30:02

Diagnostics Check		OK	00:27	
Air Blank		0.000	00:27	
Control Test		0.080	00:27	
Air Blank		0.000	00:28	
Subject Sample	# 1		00:28	Subject Test Refused
Breath Volume		0.000	00:28	
Air Blank		0.000	00:29	
Control Test		0.080	00:29	
Air Blank		0.000	00:29	
Diagnostics Check		OK	00:30	

Data Download Date/Time: 02/06/2019 12:05

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,STEVEN,D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK,STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 01/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: FISH

DATE/TIME OF TEST: 01/13/2019 00:09:55

Diagnostics Check		OK	00:03
Air Blank		0.000	00:03
Control Test		0.080	00:03
Air Blank		0.000	00:04
Subject Sample	# 1	0.271	00:05
Breath Volume		2.734	00:05
Air Blank		0.000	00:05
Air Blank		0.000	00:07
Subject Sample	# 2	0.276	00:08
Breath Volume		2.683	00:08
Air Blank		0.000	00:09
Control Test		0.080	00:09
Air Blank		0.000	00:09
Diagnostics Check		OK	00:09

Data Download Date/Time: 02/06/2019 12:05

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,BRANDON,W
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: VULETA,BEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 01/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: JIMENEZ

DATE/TIME OF TEST: 01/16/2019 02:55:50

Diagnostics Check		OK	02:49
Air Blank		0.000	02:50
Control Test		0.080	02:50
Air Blank		0.000	02:50
Subject Sample	# 1	0.122	02:51
Breath Volume		1.546	02:51
Air Blank		0.000	02:52
Air Blank		0.000	02:53
Subject Sample	# 2	0.128	02:54
Breath Volume		2.832	02:54
Air Blank		0.000	02:54
Control Test		0.080	02:55
Air Blank		0.000	02:55
Diagnostics Check		OK	02:55

Data Download Date/Time: 02/06/2019 12:05

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,STEVEN,D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK,STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 01/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 17:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:56

SUBJECT'S LAST NAME: NOLAN

DATE/TIME OF TEST: 01/19/2019 18:22:10

Diagnostics Check		OK	18:16
Air Blank		0.000	18:16
Control Test		0.080	18:16
Air Blank		0.000	18:17
Subject Sample	# 1	0.399	18:17
Breath Volume		2.578	18:17
Air Blank		0.000	18:18
Air Blank		0.000	18:20
Subject Sample	# 2	0.403	18:20
Breath Volume		2.757	18:20
Air Blank		0.000	18:21
Control Test		0.080	18:21
Air Blank		0.000	18:22
Diagnostics Check		OK	18:22

Data Download Date/Time: 02/06/2019 12:05

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMMONS,RYANE,J
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: AMMONS,RYANE
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 01/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: VANWAGNER

DATE/TIME OF TEST: 01/27/2019 03:49:20

Diagnostics Check		OK	03:46	
Air Blank		0.000	03:46	
Control Test		0.080	03:47	
Air Blank		0.000	03:47	
Subject Sample	# 1		03:47	Subject Test Refused
Breath Volume		0.000	03:47	
Air Blank		0.000	03:48	
Control Test		0.080	03:48	
Air Blank		0.000	03:49	
Diagnostics Check		OK	03:49	

Data Download Date/Time: 02/06/2019 12:05

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,BRANDON,W
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: VULETA,BEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 01/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: AKERS

DATE/TIME OF TEST: 01/29/2019 03:38:39

Diagnostics Check		OK	03:32
Air Blank		0.000	03:33
Control Test		0.079	03:33
Air Blank		0.000	03:33
Subject Sample	# 1	0.260	03:34
Breath Volume		1.777	03:34
Air Blank		0.000	03:34
Air Blank		0.000	03:36
Subject Sample	# 2	0.268	03:37
Breath Volume		2.335	03:37
Air Blank		0.000	03:37
Control Test		0.080	03:38
Air Blank		0.000	03:38
Diagnostics Check		OK	03:38

Data Download Date/Time: 02/06/2019 12:05

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,BRANDON,W
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: VULETA,BEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 01/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: VAN-VUUREN

DATE/TIME OF TEST: 01/30/2019 01:37:54

Diagnostics Check		OK	01:30
Air Blank		0.000	01:31
Control Test		0.079	01:31
Air Blank		0.000	01:31
Subject Sample	# 1	0.271	01:33
Breath Volume		2.726	01:33
Air Blank		0.000	01:33
Air Blank		0.000	01:35
Subject Sample	# 2	0.252	01:36
Breath Volume		2.089	01:36
Air Blank		0.000	01:37
Control Test		0.079	01:37
Air Blank		0.000	01:37
Diagnostics Check		OK	01:37

Data Download Date/Time: 02/06/2019 12:05

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GHOBRIAL,MARTIN,A
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: BURDICK,ROBERT
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 12/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: SEIM

DATE/TIME OF TEST: 01/01/2019 22:23:57

Diagnostics Check		OK	22:17
Air Blank		0.000	22:18
Control Test		0.079	22:18
Air Blank		0.000	22:18
Subject Sample	# 1	0.184	22:19
Breath Volume		2.824	22:19
Air Blank		0.000	22:20
Air Blank		0.000	22:21
Subject Sample	# 2	0.183	22:22
Breath Volume		2.894	22:22
Air Blank		0.000	22:23
Control Test		0.079	22:23
Air Blank		0.000	22:23
Diagnostics Check		OK	22:23

Data Download Date/Time: 01/31/2019 19:35

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,SALVADOR,
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: GARCIA,SALVADOR
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 12/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:54

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 01/03/2019 21:22:31

Diagnostics Check		OK	21:14
Air Blank		0.000	21:15
Control Test		0.079	21:15
Air Blank		0.000	21:16
Subject Sample	# 1	0.165	21:17
Breath Volume		1.398	21:17
Air Blank		0.000	21:18
Air Blank		0.000	21:20
Subject Sample	# 2	0.159	21:20
Breath Volume		2.609	21:20
Air Blank		0.000	21:21
Control Test		0.076	21:21
Air Blank		0.000	21:22
Diagnostics Check		OK	21:22

Data Download Date/Time: 01/31/2019 19:35

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TREVISOL,JOSEPH,D
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: TREVISOL,JOSEPH
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 12/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:36

SUBJECT'S LAST NAME: MCCORMACK

DATE/TIME OF TEST: 01/04/2019 19:08:49

Diagnostics Check		OK	19:02
Air Blank		0.000	19:02
Control Test		0.080	19:03
Air Blank		0.000	19:03
Subject Sample	# 1	0.285	19:04
Breath Volume		1.304	19:04
Air Blank		0.000	19:04
Air Blank		0.000	19:06
Subject Sample	# 2	0.304	19:07
Breath Volume		2.722	19:07
Air Blank		0.000	19:07
Control Test		0.078	19:08
Air Blank		0.000	19:08
Diagnostics Check		OK	19:08

Data Download Date/Time: 01/31/2019 19:35

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,MICHAEL,C
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: WILLIAMS,MICHAEL
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 12/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817C80A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: DIBARTOLOMEO

DATE/TIME OF TEST: 01/13/2019 22:35:15

Diagnostics Check		OK	22:23	
Air Blank		0.000	22:24	
Control Test		0.080	22:24	
Air Blank		0.000	22:25	
Subject Sample	# 1		22:28	No Sample Provided
Breath Volume		0.000	22:28	
Air Blank		0.000	22:28	
Air Blank		0.000	22:30	
Subject Sample	# 2		22:33	No Sample Provided
Breath Volume		0.000	22:33	
Air Blank		0.000	22:34	
Control Test		0.080	22:34	
Air Blank		0.000	22:35	
Diagnostics Check		OK	22:35	

Data Download Date/Time: 01/31/2019 19:35

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERRING,TIMOTHY,R
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: HERRING,TIMOTHY
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 12/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: BROUSSARD

DATE/TIME OF TEST: 01/19/2019 02:24:28

Diagnostics Check		OK	02:15
Air Blank		0.000	02:15
Control Test		0.079	02:16
Air Blank		0.000	02:16
Subject Sample	# 1	0.147	02:19
Breath Volume		1.292	02:19
Air Blank		0.000	02:20
Air Blank		0.000	02:21
Subject Sample	# 2	0.149	02:22
Breath Volume		1.621	02:22
Air Blank		0.000	02:23
Control Test		0.076	02:23
Air Blank		0.000	02:24
Diagnostics Check		OK	02:24

Data Download Date/Time: 01/31/2019 19:35

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANDY,KEVIN,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,BRANDON
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 12/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: NOEL

DATE/TIME OF TEST: 01/01/2019 00:30:59

Diagnostics Check		OK	00:26	
Air Blank		0.000	00:26	
Control Test		0.080	00:26	
Air Blank		0.000	00:27	
Subject Sample	# 1	0.000	00:29	
Breath Volume		1.648	00:29	
Air Blank			00:29	RFI Detect
Air Blank		0.000	00:30	
Control Test		0.079	00:30	
Air Blank		0.000	00:30	
Diagnostics Check		OK	00:30	

Data Download Date/Time: 01/04/2019 03:37

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANDY,KEVIN,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,BRANDON
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 12/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: NOEL

DATE/TIME OF TEST: 01/01/2019 00:39:17

Diagnostics Check		OK	00:32
Air Blank		0.000	00:33
Control Test		0.079	00:33
Air Blank		0.000	00:34
Subject Sample	# 1	0.000	00:34
Breath Volume		1.632	00:34
Air Blank		0.000	00:35
Air Blank		0.000	00:37
Subject Sample	# 2	0.000	00:37
Breath Volume		2.058	00:37
Air Blank		0.000	00:38
Control Test		0.079	00:38
Air Blank		0.000	00:39
Diagnostics Check		OK	00:39

Data Download Date/Time: 01/04/2019 03:37

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KANTORSKI,MARK,A
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: SAINT ELOI,ASHTON
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: FAURE

DATE/TIME OF TEST: 01/03/2019 01:36:43

Diagnostics Check		OK	01:33	
Air Blank		0.000	01:34	
Control Test		0.080	01:34	
Air Blank		0.000	01:35	
Subject Sample	# 1		01:35	Subject Test Refused
Breath Volume		0.000	01:35	
Air Blank		0.000	01:35	
Control Test		0.080	01:36	
Air Blank		0.000	01:36	
Diagnostics Check		OK	01:36	

Data Download Date/Time: 01/04/2019 03:37

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HISS,J,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,B
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 12/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 18:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: TROBAUGH

DATE/TIME OF TEST: 01/03/2019 19:35:04

Diagnostics Check		OK	19:32	
Air Blank		0.000	19:32	
Control Test		0.079	19:33	
Air Blank		0.000	19:33	
Subject Sample	# 1		19:33	Subject Test Refused
Breath Volume		0.000	19:33	
Air Blank		0.000	19:34	
Control Test		0.079	19:34	
Air Blank		0.000	19:34	
Diagnostics Check		OK	19:35	

Data Download Date/Time: 01/04/2019 03:37

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHILE,C,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: INGRAM,M
ARRESTING OFFICER AGENCY: Fellsmere PD
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 18:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: STEVENS

DATE/TIME OF TEST: 01/04/2019 19:36:51

Diagnostics Check		OK	19:30
Air Blank		0.000	19:30
Control Test		0.080	19:30
Air Blank		0.000	19:31
Subject Sample	# 1	0.209	19:32
Breath Volume		2.453	19:32
Air Blank		0.000	19:32
Air Blank		0.000	19:34
Subject Sample	# 2	0.209	19:35
Breath Volume		2.207	19:35
Air Blank		0.000	19:35
Control Test		0.079	19:36
Air Blank		0.000	19:36
Diagnostics Check		OK	19:36

Data Download Date/Time: 03/01/2019 09:34

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHILE,C,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: RADOCK,M
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 18:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: DE WITTE

DATE/TIME OF TEST: 01/04/2019 20:34:43

Diagnostics Check		OK	20:31	
Air Blank		0.000	20:32	
Control Test		0.079	20:32	
Air Blank		0.000	20:33	
Subject Sample	# 1		20:33	Subject Test Refused
Breath Volume		0.000	20:33	
Air Blank		0.000	20:33	
Control Test		0.079	20:34	
Air Blank		0.000	20:34	
Diagnostics Check		OK	20:34	

Data Download Date/Time: 03/01/2019 09:34

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANDY,KEVIN,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: HOILLET,D
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: HERBERT

DATE/TIME OF TEST: 01/05/2019 02:53:12

Diagnostics Check		OK	02:50	
Air Blank		0.000	02:50	
Control Test		0.080	02:51	
Air Blank		0.000	02:51	
Subject Sample	# 1		02:51	Subject Test Refused
Breath Volume		0.000	02:51	
Air Blank		0.000	02:52	
Control Test		0.079	02:52	
Air Blank		0.000	02:53	
Diagnostics Check		OK	02:53	

Data Download Date/Time: 03/01/2019 09:34

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLUMB,GREGORY,T
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: PLUMB,GREGORY
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:56

SUBJECT'S LAST NAME: GAVIN

DATE/TIME OF TEST: 01/05/2019 21:26:47

Diagnostics Check		OK	21:20
Air Blank		0.000	21:21
Control Test		0.079	21:21
Air Blank		0.000	21:22
Subject Sample	# 1	0.019	21:22
Breath Volume		2.558	21:22
Air Blank		0.000	21:23
Air Blank		0.000	21:24
Subject Sample	# 2	0.018	21:25
Breath Volume		1.710	21:25
Air Blank		0.000	21:25
Control Test		0.079	21:26
Air Blank		0.000	21:26
Diagnostics Check		OK	21:26

Data Download Date/Time: 03/01/2019 09:34

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FENTON,WILLIAM,B
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: FENTON,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: NEGRETE LOPEZ

DATE/TIME OF TEST: 01/06/2019 05:01:06

Diagnostics Check		OK	04:51
Air Blank		0.000	04:52
Control Test		0.080	04:52
Air Blank		0.000	04:53
Subject Sample	# 1	0.064	04:55
Breath Volume		1.222	04:55
Air Blank		0.000	04:55
Air Blank		0.000	04:57
Subject Sample	# 2	0.063	04:59
Breath Volume		1.593	04:59
Air Blank		0.000	05:00
Control Test		0.079	05:00
Air Blank		0.000	05:01
Diagnostics Check		OK	05:01

Data Download Date/Time: 03/01/2019 09:34

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANDY,KEVIN,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: FRICKE,JOANNA
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 01/07/2019 01:39:13

Diagnostics Check		OK	01:31
Air Blank		0.000	01:31
Control Test		0.079	01:32
Air Blank		0.000	01:32
Subject Sample	# 1	0.120	01:33
Breath Volume		1.195	01:33
Air Blank		0.000	01:34
Air Blank		0.000	01:36
Subject Sample	# 2	0.109	01:37
Breath Volume		1.144	01:37
Air Blank		0.000	01:38
Control Test		0.079	01:38
Air Blank		0.000	01:39
Diagnostics Check		OK	01:39

Data Download Date/Time: 03/01/2019 09:34

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANDY,KEVIN,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: BRASHEARS,ERIC
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: BECK

DATE/TIME OF TEST: 01/07/2019 02:20:29

Diagnostics Check		OK	02:13
Air Blank		0.000	02:13
Control Test		0.080	02:13
Air Blank		0.000	02:14
Subject Sample	# 1	0.166	02:15
Breath Volume		1.605	02:15
Air Blank		0.000	02:16
Air Blank		0.000	02:18
Subject Sample	# 2	0.171	02:18
Breath Volume		1.871	02:18
Air Blank		0.000	02:19
Control Test		0.078	02:19
Air Blank		0.000	02:20
Diagnostics Check		OK	02:20

Data Download Date/Time: 03/01/2019 09:35

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHILE,C,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: EVANS,Z
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 18:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: SAINT PIERRE

DATE/TIME OF TEST: 01/08/2019 19:04:27

Diagnostics Check		OK	19:00	
Air Blank		0.000	19:01	
Control Test		0.080	19:01	
Air Blank		0.000	19:02	
Subject Sample	# 1		19:03	Subject Test Refused
Breath Volume		0.000	19:03	
Air Blank		0.000	19:03	
Control Test		0.079	19:03	
Air Blank		0.000	19:04	
Diagnostics Check		OK	19:04	

Data Download Date/Time: 03/01/2019 09:35

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHILE,C,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,B
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 20:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: POOLE

DATE/TIME OF TEST: 01/08/2019 21:46:02

Diagnostics Check		OK	21:37	
Air Blank		0.000	21:38	
Control Test		0.080	21:38	
Air Blank		0.000	21:38	
Subject Sample	# 1		21:42	No Sample Provided
Breath Volume		0.000	21:42	
Air Blank		0.000	21:42	
Air Blank		0.000	21:44	
Subject Sample	# 2		21:44	Subject Test Refused
Breath Volume		0.000	21:44	
Air Blank		0.000	21:45	
Control Test		0.079	21:45	
Air Blank		0.000	21:45	
Diagnostics Check		OK	21:46	

Data Download Date/Time: 03/01/2019 09:35

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLIDAY,LATONYA,A
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: HOILETT,DWAYNE
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 01/12/2019 01:30:03

Diagnostics Check		OK	01:23
Air Blank		0.000	01:23
Control Test		0.080	01:24
Air Blank		0.000	01:24
Subject Sample	# 1	0.020	01:25
Breath Volume		2.191	01:25
Air Blank		0.000	01:26
Air Blank		0.000	01:28
Subject Sample	# 2	0.018	01:28
Breath Volume		1.664	01:28
Air Blank		0.000	01:29
Control Test		0.080	01:29
Air Blank		0.000	01:29
Diagnostics Check		OK	01:30

Data Download Date/Time: 03/01/2019 09:35

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FONSECA,GIOVANNY,
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: FONSECA,GIOVANNY
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: FLYNN

DATE/TIME OF TEST: 01/12/2019 05:32:09

Diagnostics Check		OK	05:25
Air Blank		0.000	05:26
Control Test		0.079	05:26
Air Blank		0.000	05:26
Subject Sample	# 1	0.257	05:27
Breath Volume		2.308	05:27
Air Blank		0.000	05:28
Air Blank		0.000	05:29
Subject Sample	# 2	0.256	05:30
Breath Volume		2.488	05:30
Air Blank		0.000	05:31
Control Test		0.079	05:31
Air Blank		0.000	05:32
Diagnostics Check		OK	05:32

Data Download Date/Time: 03/01/2019 09:35

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MAEGAN,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: JAWORSKI,KEVIN
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 19:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 01/12/2019 20:45:58

Diagnostics Check		OK	20:37
Air Blank		0.000	20:37
Control Test		0.079	20:38
Air Blank		0.000	20:38
Subject Sample	# 1	0.262	20:40
Breath Volume		1.589	20:40
Air Blank		0.000	20:40
Air Blank		0.000	20:42
Subject Sample	# 2	0.261	20:44
Breath Volume		1.710	20:44
Air Blank		0.000	20:45
Control Test		0.078	20:45
Air Blank		0.000	20:45
Diagnostics Check		OK	20:45

Data Download Date/Time: 03/01/2019 09:35

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DEPUTY,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: SAINT ELOI,A
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 20:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: HERNANDEZ-HERRERA

DATE/TIME OF TEST: 01/12/2019 21:59:07

Diagnostics Check		OK	21:56	
Air Blank		0.000	21:56	
Control Test		0.080	21:56	
Air Blank		0.000	21:57	
Subject Sample	# 1		21:57	Subject Test Refused
Breath Volume		0.000	21:57	
Air Blank		0.000	21:58	
Control Test		0.079	21:58	
Air Blank		0.000	21:59	
Diagnostics Check		OK	21:59	

Data Download Date/Time: 03/01/2019 09:35

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MAEGAN,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: HOILETT,DEPUTY
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 01/12/2019 22:25:51

Diagnostics Check		OK	22:21	
Air Blank		0.000	22:22	
Control Test		0.079	22:22	
Air Blank		0.000	22:22	
Subject Sample	# 1		22:24	Subject Test Refused
Breath Volume		0.000	22:24	
Air Blank		0.000	22:24	
Control Test		0.080	22:25	
Air Blank		0.000	22:25	
Diagnostics Check		OK	22:25	

Data Download Date/Time: 03/01/2019 09:35

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DEPUTY,T
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: LE,Q
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 01/12/2019 23:09:57

Diagnostics Check		OK	23:07	
Air Blank		0.000	23:07	
Control Test		0.080	23:07	
Air Blank		0.000	23:08	
Subject Sample	# 1		23:08	Subject Test Refused
Breath Volume		0.000	23:08	
Air Blank		0.000	23:09	
Control Test		0.079	23:09	
Air Blank		0.000	23:09	
Diagnostics Check		OK	23:09	

Data Download Date/Time: 03/01/2019 09:35

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUDOLPH,B,D
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,B
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 18:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: KUNO

DATE/TIME OF TEST: 01/13/2019 19:41:58

Diagnostics Check		OK	19:39	
Air Blank		0.000	19:39	
Control Test		0.080	19:39	
Air Blank		0.000	19:40	
Subject Sample	# 1		19:40	Subject Test Refused
Breath Volume		0.000	19:40	
Air Blank		0.000	19:41	
Control Test		0.079	19:41	
Air Blank		0.000	19:41	
Diagnostics Check		OK	19:41	

Data Download Date/Time: 03/01/2019 09:35

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WLLIAMS,QUINTON,T
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: FLORES,DAVID
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: TELLEZ RIOS

DATE/TIME OF TEST: 01/15/2019 21:58:08

Diagnostics Check		OK	21:51
Air Blank		0.000	21:52
Control Test		0.081	21:52
Air Blank		0.000	21:52
Subject Sample	# 1	0.238	21:53
Breath Volume		2.273	21:53
Air Blank		0.000	21:54
Air Blank		0.000	21:56
Subject Sample	# 2	0.248	21:56
Breath Volume		2.613	21:56
Air Blank		0.000	21:57
Control Test		0.080	21:57
Air Blank		0.000	21:58
Diagnostics Check		OK	21:58

Data Download Date/Time: 03/01/2019 09:35

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWINDELL,CHRISTOPHER,E
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: INGRAM,JENNIFER
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: BUCHER

DATE/TIME OF TEST: 01/18/2019 00:49:47

Diagnostics Check		OK	00:42
Air Blank		0.000	00:42
Control Test		0.079	00:43
Air Blank		0.000	00:43
Subject Sample	# 1	0.163	00:45
Breath Volume		2.339	00:45
Air Blank		0.000	00:45
Air Blank		0.000	00:47
Subject Sample	# 2	0.159	00:48
Breath Volume		2.109	00:48
Air Blank		0.000	00:48
Control Test		0.080	00:49
Air Blank		0.000	00:49
Diagnostics Check		OK	00:49

Data Download Date/Time: 03/01/2019 09:35

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHILE,C,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: POLLACEK,J
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 20:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: BYNUM

DATE/TIME OF TEST: 01/18/2019 21:12:04

Diagnostics Check		OK	21:05
Air Blank		0.000	21:06
Control Test		0.079	21:06
Air Blank		0.000	21:07
Subject Sample	# 1	0.203	21:07
Breath Volume		2.777	21:07
Air Blank		0.000	21:08
Air Blank		0.000	21:10
Subject Sample	# 2	0.192	21:10
Breath Volume		2.285	21:10
Air Blank		0.000	21:11
Control Test		0.079	21:11
Air Blank		0.000	21:11
Diagnostics Check		OK	21:12

Data Download Date/Time: 03/01/2019 09:35

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KUEHN,KYLE,B
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: AYALA,KASSANDRA
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: BIEBER

DATE/TIME OF TEST: 01/18/2019 23:31:13

Diagnostics Check		OK	23:28	
Air Blank		0.000	23:28	
Control Test		0.080	23:29	
Air Blank		0.000	23:29	
Subject Sample	# 1		23:29	Subject Test Refused
Breath Volume		0.000	23:29	
Air Blank		0.000	23:30	
Control Test		0.079	23:30	
Air Blank		0.000	23:31	
Diagnostics Check		OK	23:31	

Data Download Date/Time: 03/01/2019 09:35

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUDOLPH,B,D
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: POLLACEK,J
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 01/19/2019 01:31:55

Diagnostics Check		OK	01:25
Air Blank		0.000	01:26
Control Test		0.079	01:26
Air Blank		0.000	01:27
Subject Sample	# 1	0.101	01:27
Breath Volume		1.703	01:27
Air Blank		0.000	01:28
Air Blank		0.000	01:29
Subject Sample	# 2	0.098	01:30
Breath Volume		1.285	01:30
Air Blank		0.000	01:31
Control Test		0.079	01:31
Air Blank		0.000	01:31
Diagnostics Check		OK	01:31

Data Download Date/Time: 03/01/2019 09:35

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LACHOWICZ,RUSSELL,M
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: ORTIZ,TAYLOR
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 14:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:22

SUBJECT'S LAST NAME: PROFITT

DATE/TIME OF TEST: 01/21/2019 14:52:11

Diagnostics Check		OK	14:46
Air Blank		0.000	14:46
Control Test		0.080	14:46
Air Blank		0.000	14:47
Subject Sample	# 1	0.179	14:47
Breath Volume		1.882	14:47
Air Blank		0.000	14:48
Air Blank		0.000	14:50
Subject Sample	# 2	0.180	14:50
Breath Volume		2.160	14:50
Air Blank		0.000	14:51
Control Test		0.079	14:51
Air Blank		0.000	14:52
Diagnostics Check		OK	14:52

Data Download Date/Time: 03/01/2019 09:35

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLIDAY,LATONYA,A
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,BRANDON
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: WEBB

DATE/TIME OF TEST: 01/22/2019 03:11:47

Diagnostics Check		OK	03:05
Air Blank		0.000	03:06
Control Test		0.080	03:06
Air Blank		0.000	03:06
Subject Sample	# 1	0.000	03:07
Breath Volume		1.117	03:07
Air Blank		0.000	03:08
Air Blank		0.000	03:10
Subject Sample	# 2	0.000	03:10
Breath Volume		1.585	03:10
Air Blank		0.000	03:10
Control Test		0.080	03:11
Air Blank		0.000	03:11
Diagnostics Check		OK	03:11

Data Download Date/Time: 03/01/2019 09:35

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,QUINTON,T
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: FLORES,DEPUTY
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 01/23/2019 22:03:43

Diagnostics Check		OK	21:57	
Air Blank		0.000	21:57	
Control Test		0.080	21:58	
Air Blank		0.000	21:58	
Subject Sample	# 1	0.372	21:59	
Breath Volume		3.160	21:59	
Air Blank		0.000	21:59	
Air Blank		0.000	22:01	
Subject Sample	# 2		22:02	Slope Not Met
Breath Volume		1.644	22:02	
Air Blank		0.000	22:02	
Control Test		0.078	22:03	
Air Blank		0.000	22:03	
Diagnostics Check		OK	22:03	

Data Download Date/Time: 03/01/2019 09:35

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,QUINTON,T
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: FLORES,DEPUTY
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:07

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 01/23/2019 22:33:02

Diagnostics Check		OK	22:26
Air Blank		0.000	22:27
Control Test		0.078	22:27
Air Blank		0.000	22:28
Subject Sample	# 1	0.368	22:28
Breath Volume		3.468	22:28
Air Blank		0.000	22:29
Air Blank		0.000	22:30
Subject Sample	# 2	0.362	22:31
Breath Volume		2.558	22:31
Air Blank		0.000	22:32
Control Test		0.079	22:32
Air Blank		0.000	22:32
Diagnostics Check		OK	22:33

Data Download Date/Time: 03/01/2019 09:35

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLIDAY,LATONYA,A
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: DELEON,TEA
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: STRIBBLING

DATE/TIME OF TEST: 01/25/2019 02:32:25

Diagnostics Check		OK	02:21	
Air Blank		0.000	02:22	
Control Test		0.079	02:22	
Air Blank		0.000	02:22	
Subject Sample	# 1	0.116	02:26	Volume Not Met
Breath Volume		0.605	02:26	
Air Blank		0.000	02:26	
Air Blank		0.000	02:28	
Subject Sample	# 2		02:30	Subject Test Refused
Breath Volume		0.503	02:30	
Air Blank		0.000	02:31	
Control Test		0.079	02:31	
Air Blank		0.000	02:32	
Diagnostics Check		OK	02:32	

Data Download Date/Time: 03/01/2019 09:35

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,QUINTON,T
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: POLLACEK,J
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: SEARS

DATE/TIME OF TEST: 01/25/2019 21:59:32

Diagnostics Check		OK	21:52
Air Blank		0.000	21:53
Control Test		0.080	21:53
Air Blank		0.000	21:54
Subject Sample	# 1	0.223	21:55
Breath Volume		1.257	21:55
Air Blank		0.000	21:55
Air Blank		0.000	21:57
Subject Sample	# 2	0.233	21:58
Breath Volume		1.109	21:58
Air Blank		0.000	21:58
Control Test		0.080	21:58
Air Blank		0.000	21:59
Diagnostics Check		OK	21:59

Data Download Date/Time: 03/01/2019 09:35

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,QUINTON,T
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: FARINACCI,.
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 20:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: CASTELLANO

DATE/TIME OF TEST: 01/25/2019 23:14:53

Diagnostics Check		OK	23:08
Air Blank		0.000	23:09
Control Test		0.080	23:09
Air Blank		0.000	23:10
Subject Sample	# 1	0.109	23:10
Breath Volume		2.210	23:10
Air Blank		0.000	23:11
Air Blank		0.000	23:13
Subject Sample	# 2	0.109	23:13
Breath Volume		3.000	23:13
Air Blank		0.000	23:13
Control Test		0.080	23:14
Air Blank		0.000	23:14
Diagnostics Check		OK	23:14

Data Download Date/Time: 03/01/2019 09:35

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUDOLPH,B,D
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: EGGERS,R
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 01/26/2019 00:56:41

Diagnostics Check		OK	00:53	
Air Blank		0.000	00:54	
Control Test		0.080	00:54	
Air Blank		0.000	00:55	
Subject Sample	# 1		00:55	Subject Test Refused
Breath Volume		0.000	00:55	
Air Blank		0.000	00:55	
Control Test		0.080	00:56	
Air Blank		0.000	00:56	
Diagnostics Check		OK	00:56	

Data Download Date/Time: 03/01/2019 09:35

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUDOLPH,B,D
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,B
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: KRAUSE

DATE/TIME OF TEST: 01/26/2019 01:23:46

Diagnostics Check		OK	01:17
Air Blank		0.000	01:17
Control Test		0.080	01:18
Air Blank		0.000	01:18
Subject Sample	# 1	0.048	01:19
Breath Volume		1.742	01:19
Air Blank		0.000	01:19
Air Blank		0.000	01:21
Subject Sample	# 2	0.052	01:22
Breath Volume		1.699	01:22
Air Blank		0.000	01:22
Control Test		0.080	01:23
Air Blank		0.000	01:23
Diagnostics Check		OK	01:23

Data Download Date/Time: 03/01/2019 09:35

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MAEGAN,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: GAST,E
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 18:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: COLVIN

DATE/TIME OF TEST: 01/26/2019 20:21:06

Diagnostics Check		OK	20:18	
Air Blank		0.000	20:18	
Control Test		0.080	20:19	
Air Blank		0.000	20:19	
Subject Sample	# 1		20:19	Subject Test Refused
Breath Volume		0.000	20:19	
Air Blank		0.000	20:20	
Control Test		0.080	20:20	
Air Blank		0.000	20:21	
Diagnostics Check		OK	20:21	

Data Download Date/Time: 03/01/2019 09:35

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MAEGAN,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: BARARCO,B
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 01/26/2019 23:28:47

Diagnostics Check		OK	23:22
Air Blank		0.000	23:23
Control Test		0.079	23:23
Air Blank		0.000	23:23
Subject Sample	# 1	0.219	23:24
Breath Volume		1.839	23:24
Air Blank		0.000	23:24
Air Blank		0.000	23:26
Subject Sample	# 2	0.225	23:27
Breath Volume		1.425	23:27
Air Blank		0.000	23:27
Control Test		0.079	23:28
Air Blank		0.000	23:28
Diagnostics Check		OK	23:28

Data Download Date/Time: 03/01/2019 09:35

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUDOLPH,B,D
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,B
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: SCOZZARI

DATE/TIME OF TEST: 01/27/2019 03:41:34

Diagnostics Check		OK	03:38	
Air Blank		0.000	03:39	
Control Test		0.080	03:39	
Air Blank		0.000	03:40	
Subject Sample	# 1		03:40	Subject Test Refused
Breath Volume		0.000	03:40	
Air Blank		0.000	03:40	
Control Test		0.080	03:40	
Air Blank		0.000	03:41	
Diagnostics Check		OK	03:41	

Data Download Date/Time: 03/01/2019 09:35

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MAEGAN,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: CASTANO,C
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 20:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 01/29/2019 22:01:27

Diagnostics Check		OK	21:48	
Air Blank		0.000	21:49	
Control Test		0.079	21:49	
Air Blank		0.000	21:49	
Subject Sample	# 1	0.164	21:53	Volume Not Met
Breath Volume		1.097	21:53	
Air Blank		0.000	21:53	
Air Blank		0.000	21:55	
Subject Sample	# 2	0.169	21:56	No .020 agreement
Breath Volume		1.664	21:56	
Air Blank		0.000	21:57	
Air Blank		0.000	21:59	
Subject Sample	# 3	0.192	21:59	No .020 agreement
Breath Volume		1.429	21:59	
Air Blank		0.000	22:00	
Control Test		0.079	22:00	
Air Blank		0.000	22:01	
Diagnostics Check		OK	22:01	

Data Download Date/Time: 03/01/2019 09:35

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANDY,KEVIN,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: KAPLAN,NICHOLAS
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: REYNOLDS

DATE/TIME OF TEST: 01/04/2019 03:16:53

Diagnostics Check		OK	03:08
Air Blank		0.000	03:08
Control Test		0.081	03:09
Air Blank		0.000	03:09
Subject Sample	# 1	0.103	03:11
Breath Volume		1.128	03:11
Air Blank		0.000	03:12
Air Blank		0.000	03:14
Subject Sample	# 2	0.103	03:15
Breath Volume		1.511	03:15
Air Blank		0.000	03:15
Control Test		0.080	03:16
Air Blank		0.000	03:16
Diagnostics Check		OK	03:16

Data Download Date/Time: 03/01/2019 09:44

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAYA,ALFREDO,N
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: VARGAS,MARIA
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 12/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: REAVES

DATE/TIME OF TEST: 01/20/2019 21:33:24

Diagnostics Check		OK	21:21	
Air Blank		0.000	21:22	
Control Test		0.081	21:22	
Air Blank		0.000	21:23	
Subject Sample	# 1	0.127	21:26	Volume Not Met
Breath Volume		0.058	21:26	
Air Blank		0.000	21:26	
Air Blank		0.000	21:28	
Subject Sample	# 2	0.105	21:31	Volume Not Met
Breath Volume		0.574	21:31	
Air Blank		0.000	21:32	
Control Test		0.081	21:32	
Air Blank		0.000	21:33	
Diagnostics Check		OK	21:33	

Data Download Date/Time: 01/31/2019 15:49

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMBRIA,STEVEN,J
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: CAMBRIA,STEVEN
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 12/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 01/01/2019 01:28:55

Diagnostics Check		OK	01:24	
Air Blank		0.000	01:24	
Control Test		0.080	01:25	
Air Blank		0.000	01:25	
Subject Sample	# 1		01:28	Sequence Aborted
Breath Volume		0.000	01:28	
Air Blank		0.000	01:28	

Data Download Date/Time: 01/30/2019 07:50

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMBRIA,STEVEN,J
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: MILLER,R
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: GUZMAN

DATE/TIME OF TEST: 01/01/2019 04:09:46

Diagnostics Check		OK	04:03
Air Blank		0.000	04:03
Control Test		0.080	04:04
Air Blank		0.000	04:04
Subject Sample	# 1	0.210	04:05
Breath Volume		2.128	04:05
Air Blank		0.000	04:06
Air Blank		0.000	04:07
Subject Sample	# 2	0.210	04:08
Breath Volume		2.472	04:08
Air Blank		0.000	04:08
Control Test		0.078	04:09
Air Blank		0.000	04:09
Diagnostics Check		OK	04:09

Data Download Date/Time: 01/30/2019 07:50

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMBRIA,STEVEN,J
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: CAMBRIA,STEVEN
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 12/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 01/01/2019 21:39:33

Diagnostics Check		OK	21:32
Air Blank		0.000	21:33
Control Test		0.081	21:33
Air Blank		0.000	21:34
Subject Sample	# 1	0.162	21:34
Breath Volume		2.015	21:34
Air Blank		0.000	21:35
Air Blank		0.000	21:37
Subject Sample	# 2	0.164	21:37
Breath Volume		3.410	21:37
Air Blank		0.000	21:38
Control Test		0.079	21:38
Air Blank		0.000	21:39
Diagnostics Check		OK	21:39

Data Download Date/Time: 01/30/2019 07:50

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NICK,A.,
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: NICK,A.
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 12/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: ACOSTA

DATE/TIME OF TEST: 01/01/2019 23:34:57

Diagnostics Check		OK	23:28
Air Blank		0.000	23:28
Control Test		0.080	23:29
Air Blank		0.000	23:29
Subject Sample	# 1	0.000	23:30
Breath Volume		4.156	23:30
Air Blank		0.000	23:30
Air Blank		0.000	23:32
Subject Sample	# 2	0.000	23:33
Breath Volume		3.472	23:33
Air Blank		0.000	23:34
Control Test		0.079	23:34
Air Blank		0.000	23:34
Diagnostics Check		OK	23:34

Data Download Date/Time: 01/30/2019 07:50

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,NICKKI,N
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FOLADVANDI,M
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 12/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: FINCH

DATE/TIME OF TEST: 01/02/2019 00:26:13

Diagnostics Check		OK	00:19
Air Blank		0.000	00:19
Control Test		0.081	00:20
Air Blank		0.000	00:20
Subject Sample	# 1	0.192	00:21
Breath Volume		2.187	00:21
Air Blank		0.000	00:22
Air Blank		0.000	00:23
Subject Sample	# 2	0.180	00:24
Breath Volume		2.277	00:24
Air Blank		0.000	00:25
Control Test		0.080	00:25
Air Blank		0.000	00:26
Diagnostics Check		OK	00:26

Data Download Date/Time: 01/30/2019 07:50

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,NICKKI,N
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: TOMLINSON,E
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 12/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: MCKINNIE

DATE/TIME OF TEST: 01/05/2019 00:38:41

Diagnostics Check		OK	00:24	
Air Blank		0.000	00:24	
Control Test		0.080	00:25	
Air Blank		0.000	00:25	
Subject Sample	# 1	0.184	00:28	
Breath Volume		2.953	00:28	
Air Blank		0.000	00:28	
Air Blank		0.000	00:30	
Subject Sample	# 2	0.150	00:32	
Breath Volume		2.476	00:32	
Air Blank		0.000	00:33	
Air Blank		0.000	00:34	
Subject Sample	# 3	0.164	00:37	
Breath Volume		3.144	00:37	
Air Blank		0.000	00:37	
Control Test		0.073	00:38	Control Outside Tolerance
Air Blank		0.000	00:38	

Data Download Date/Time: 01/30/2019 07:50

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,NICKKI,N
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: TOMLINSON,E
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 12/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: MCKINNIE

DATE/TIME OF TEST: 01/05/2019 00:51:51

Diagnostics Check		OK	00:45
Air Blank		0.000	00:46
Control Test		0.077	00:46
Air Blank		0.000	00:46
Subject Sample	# 1	0.161	00:47
Breath Volume		1.730	00:47
Air Blank		0.000	00:47
Air Blank		0.000	00:49
Subject Sample	# 2	0.166	00:50
Breath Volume		2.238	00:50
Air Blank		0.000	00:50
Control Test		0.079	00:51
Air Blank		0.000	00:51
Diagnostics Check		OK	00:51

Data Download Date/Time: 01/30/2019 07:50

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,DANIEL,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MILLER,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:56

SUBJECT'S LAST NAME: DANIEL

DATE/TIME OF TEST: 01/05/2019 05:37:49

Diagnostics Check		OK	05:31
Air Blank		0.000	05:31
Control Test		0.079	05:32
Air Blank		0.000	05:32
Subject Sample	# 1	0.136	05:33
Breath Volume		2.449	05:33
Air Blank		0.000	05:33
Air Blank		0.000	05:35
Subject Sample	# 2	0.128	05:36
Breath Volume		2.261	05:36
Air Blank		0.000	05:36
Control Test		0.079	05:37
Air Blank		0.000	05:37
Diagnostics Check		OK	05:37

Data Download Date/Time: 01/30/2019 07:50

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEDOREK,JEFFREY,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FELICIEN,OFC
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 12/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: ALDRIDGE

DATE/TIME OF TEST: 01/06/2019 04:10:02

Diagnostics Check		OK	04:01	
Air Blank		0.000	04:01	
Control Test		0.080	04:01	
Air Blank		0.000	04:02	
Subject Sample	# 1	0.137	04:02	
Breath Volume		1.589	04:02	
Air Blank		0.000	04:03	
Air Blank		0.000	04:05	
Subject Sample	# 2	0.138	04:08	Volume Not Met
Breath Volume		0.460	04:08	
Air Blank			04:09	Purge Fail
Air Blank			04:10	Purge Fail

Data Download Date/Time: 01/30/2019 07:50

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEDOREK,J,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FELICIEN,OFC
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 12/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: ALDRIDGE

DATE/TIME OF TEST: 01/06/2019 04:53:02

Diagnostics Check		OK	04:45
Air Blank		0.000	04:46
Control Test		0.079	04:46
Air Blank		0.000	04:47
Subject Sample	# 1	0.128	04:47
Breath Volume		1.683	04:47
Air Blank		0.000	04:48
Air Blank		0.000	04:50
Subject Sample	# 2	0.126	04:51
Breath Volume		1.804	04:51
Air Blank		0.000	04:52
Control Test		0.078	04:52
Air Blank		0.000	04:52
Diagnostics Check		OK	04:53

Data Download Date/Time: 01/30/2019 07:50

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUBRANO,DEPUTY,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MILLER,DEPUTY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:36

SUBJECT'S LAST NAME: HOSTETLER

DATE/TIME OF TEST: 01/06/2019 20:09:52

Diagnostics Check		OK	20:03
Air Blank		0.000	20:03
Control Test		0.080	20:03
Air Blank		0.000	20:04
Subject Sample	# 1	0.155	20:05
Breath Volume		1.773	20:05
Air Blank		0.000	20:05
Air Blank		0.000	20:07
Subject Sample	# 2	0.145	20:08
Breath Volume		1.589	20:08
Air Blank		0.000	20:08
Control Test		0.080	20:09
Air Blank		0.000	20:09
Diagnostics Check		OK	20:09

Data Download Date/Time: 01/30/2019 07:50

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUBRANO,DEPUTY,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEPUTY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: SOKOLOW

DATE/TIME OF TEST: 01/06/2019 20:50:03

Diagnostics Check		OK	20:39	
Air Blank		0.000	20:39	
Control Test		0.081	20:40	
Air Blank		0.000	20:40	
Subject Sample	# 1	0.139	20:43	Volume Not Met
Breath Volume		0.582	20:43	
Air Blank		0.000	20:44	
Air Blank		0.000	20:46	
Subject Sample	# 2		20:48	Subject Test Refused
Breath Volume		0.195	20:48	
Air Blank		0.000	20:49	
Control Test		0.079	20:49	
Air Blank		0.000	20:49	
Diagnostics Check		OK	20:50	

Data Download Date/Time: 01/30/2019 07:50

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,DANIEL,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: OSTROSKI,OFC.
ARRESTING OFFICER AGENCY: Venice PD
LAST AGENCY INSPECTION: 12/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: STEARNS

DATE/TIME OF TEST: 01/09/2019 22:13:59

Diagnostics Check		OK	22:10	
Air Blank		0.000	22:10	
Control Test		0.076	22:11	
Air Blank		0.000	22:11	
Subject Sample	# 1		22:12	Subject Test Refused
Breath Volume		0.000	22:12	
Air Blank		0.000	22:13	
Control Test		0.081	22:13	
Air Blank		0.000	22:13	
Diagnostics Check		OK	22:13	

Data Download Date/Time: 01/30/2019 07:50

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOETGER,ELI,L
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: LISSOW,SHAUN
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 12/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:22

SUBJECT'S LAST NAME: QUINN

DATE/TIME OF TEST: 01/13/2019 19:57:25

Diagnostics Check		OK	19:51
Air Blank		0.000	19:51
Control Test		0.080	19:52
Air Blank		0.000	19:52
Subject Sample	# 1	0.151	19:52
Breath Volume		1.742	19:52
Air Blank		0.000	19:53
Air Blank		0.000	19:55
Subject Sample	# 2	0.152	19:55
Breath Volume		2.449	19:55
Air Blank		0.000	19:56
Control Test		0.079	19:56
Air Blank		0.000	19:57
Diagnostics Check		OK	19:57

Data Download Date/Time: 01/30/2019 07:50

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERSON,STEFANIE,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: NICHOLS,CAMERON
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: MCCARTHY

DATE/TIME OF TEST: 01/14/2019 00:40:07

Diagnostics Check		OK	00:33
Air Blank		0.000	00:34
Control Test		0.080	00:34
Air Blank		0.000	00:34
Subject Sample	# 1	0.299	00:35
Breath Volume		2.765	00:35
Air Blank		0.000	00:36
Air Blank		0.000	00:37
Subject Sample	# 2	0.303	00:38
Breath Volume		2.792	00:38
Air Blank		0.000	00:39
Control Test		0.079	00:39
Air Blank		0.000	00:40
Diagnostics Check		OK	00:40

Data Download Date/Time: 01/30/2019 07:50

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSTYAL,MATT,A
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KOSTYAL,MATT
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: SCHEINGOLD

DATE/TIME OF TEST: 01/16/2019 22:39:42

Diagnostics Check		OK	22:30	
Air Blank		0.000	22:31	
Control Test		0.081	22:31	
Air Blank		0.000	22:31	
Subject Sample	# 1	0.163	22:35	Volume Not Met
Breath Volume		0.355	22:35	
Air Blank		0.000	22:35	
Air Blank		0.000	22:37	
Subject Sample	# 2		22:38	Interferent Detect
Breath Volume		1.175	22:38	
Air Blank		0.000	22:38	
Control Test		0.078	22:39	
Air Blank		0.000	22:39	
Diagnostics Check		OK	22:39	

Data Download Date/Time: 01/30/2019 07:50

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSTYAL,MATT,A
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KOSTYAL,MATT
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: SCHEINGOLD

DATE/TIME OF TEST: 01/16/2019 22:53:32

Diagnostics Check		OK	22:41	
Air Blank		0.000	22:42	
Control Test		0.079	22:42	
Air Blank		0.000	22:42	
Subject Sample	# 1	0.185	22:46	Volume Not Met
Breath Volume		0.789	22:46	
Air Blank		0.000	22:46	
Air Blank		0.000	22:48	
Subject Sample	# 2	0.205	22:51	Volume Not Met
Breath Volume		0.804	22:51	
Air Blank		0.000	22:52	
Control Test		0.078	22:52	
Air Blank		0.000	22:53	
Diagnostics Check		OK	22:53	

Data Download Date/Time: 01/30/2019 07:50

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMBRIA,STEVEN,J
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: CAMBRIA,STEVEN
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 12/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:04

SUBJECT'S LAST NAME: GRANT

DATE/TIME OF TEST: 01/17/2019 19:37:03

Diagnostics Check		OK	19:30
Air Blank		0.000	19:30
Control Test		0.080	19:31
Air Blank		0.000	19:31
Subject Sample	# 1	0.100	19:32
Breath Volume		2.375	19:32
Air Blank		0.000	19:33
Air Blank		0.000	19:34
Subject Sample	# 2	0.095	19:35
Breath Volume		2.257	19:35
Air Blank		0.000	19:36
Control Test		0.078	19:36
Air Blank		0.000	19:36
Diagnostics Check		OK	19:37

Data Download Date/Time: 01/30/2019 07:50

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VERMILLION,OFFICER,B
BREATH TEST OPERATOR AGENCY: Sarasota PD
ARRESTING OFFICER: VERMILLION,OFFICER
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 12/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: CHAVEZ-RAMIREZ

DATE/TIME OF TEST: 01/19/2019 03:27:30

Diagnostics Check		OK	03:15	
Air Blank		0.000	03:16	
Control Test		0.081	03:16	
Air Blank		0.000	03:17	
Subject Sample	# 1	0.169	03:20	Volume Not Met
Breath Volume		0.062	03:20	
Air Blank		0.000	03:21	
Air Blank		0.000	03:22	
Subject Sample	# 2		03:25	No Sample Provided
Breath Volume		0.000	03:25	
Air Blank		0.000	03:26	
Control Test		0.077	03:26	
Air Blank		0.000	03:27	
Diagnostics Check		OK	03:27	

Data Download Date/Time: 01/30/2019 07:50

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VERMILLION,OFFICER,B
BREATH TEST OPERATOR AGENCY: Sarasota PD
ARRESTING OFFICER: VERMILLION,OFFICER
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 12/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: CHAVEZ-RAMIREZ

DATE/TIME OF TEST: 01/19/2019 03:44:51

Diagnostics Check		OK	03:33	
Air Blank		0.000	03:34	
Control Test		0.076	03:34	
Air Blank		0.000	03:34	
Subject Sample	# 1		03:38	No Sample Provided
Breath Volume		0.000	03:38	
Air Blank		0.000	03:38	
Air Blank		0.000	03:40	
Subject Sample	# 2		03:43	Subject Test Refused
Breath Volume		0.000	03:43	
Air Blank		0.000	03:43	
Control Test		0.080	03:44	
Air Blank		0.000	03:44	
Diagnostics Check		OK	03:44	

Data Download Date/Time: 01/30/2019 07:50

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,DANIEL,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: TAYLOR,OFC
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 12/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: LEWANDER

DATE/TIME OF TEST: 01/23/2019 19:39:29

Diagnostics Check		OK	19:33
Air Blank		0.000	19:33
Control Test		0.080	19:34
Air Blank		0.000	19:34
Subject Sample	# 1	0.189	19:35
Breath Volume		1.765	19:35
Air Blank		0.000	19:35
Air Blank		0.000	19:37
Subject Sample	# 2	0.193	19:37
Breath Volume		2.230	19:37
Air Blank		0.000	19:38
Control Test		0.079	19:38
Air Blank		0.000	19:39
Diagnostics Check		OK	19:39

Data Download Date/Time: 01/30/2019 07:50

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NICK,A.,
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: NICK,A.
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 12/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 14:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:50

SUBJECT'S LAST NAME: GREENE

DATE/TIME OF TEST: 01/24/2019 16:25:57

Diagnostics Check		OK	16:19
Air Blank		0.000	16:20
Control Test		0.080	16:20
Air Blank		0.000	16:21
Subject Sample	# 1	0.000	16:21
Breath Volume		4.312	16:21
Air Blank		0.000	16:22
Air Blank		0.000	16:24
Subject Sample	# 2	0.000	16:24
Breath Volume		2.300	16:24
Air Blank		0.000	16:25
Control Test		0.079	16:25
Air Blank		0.000	16:25
Diagnostics Check		OK	16:25

Data Download Date/Time: 01/30/2019 07:50

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMBRIA,STEVEN,J
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: CAMBRIA,STEVEN
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 12/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: KOLOGINSKI

DATE/TIME OF TEST: 01/25/2019 19:57:29

Diagnostics Check		OK	19:51
Air Blank		0.000	19:51
Control Test		0.082	19:52
Air Blank		0.000	19:52
Subject Sample	# 1	0.000	19:53
Breath Volume		3.167	19:53
Air Blank		0.000	19:53
Air Blank		0.000	19:55
Subject Sample	# 2	0.000	19:56
Breath Volume		3.089	19:56
Air Blank		0.000	19:56
Control Test		0.080	19:56
Air Blank		0.000	19:57
Diagnostics Check		OK	19:57

Data Download Date/Time: 01/30/2019 07:50

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERKINS,OFC,W
BREATH TEST OPERATOR AGENCY: Sarasota PD
ARRESTING OFFICER: PERKINS,OFC
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 12/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: TORMEY

DATE/TIME OF TEST: 01/26/2019 00:52:53

Diagnostics Check		OK	00:45
Air Blank		0.000	00:46
Control Test		0.080	00:46
Air Blank		0.000	00:47
Subject Sample	# 1	0.178	00:48
Breath Volume		1.472	00:48
Air Blank		0.000	00:48
Air Blank		0.000	00:50
Subject Sample	# 2	0.175	00:51
Breath Volume		3.675	00:51
Air Blank		0.000	00:51
Control Test		0.080	00:52
Air Blank		0.000	00:52
Diagnostics Check		OK	00:52

Data Download Date/Time: 01/30/2019 07:50

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMBRIA,STEVEN,J
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: CAMBRIA,STEVEN
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 12/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: DARDZINSKI

DATE/TIME OF TEST: 01/26/2019 02:03:15

Diagnostics Check		OK	01:57
Air Blank		0.000	01:57
Control Test		0.081	01:57
Air Blank		0.000	01:58
Subject Sample	# 1	0.000	01:58
Breath Volume		3.066	01:58
Air Blank		0.000	01:59
Air Blank		0.000	02:01
Subject Sample	# 2	0.000	02:01
Breath Volume		1.519	02:01
Air Blank		0.000	02:02
Control Test		0.080	02:02
Air Blank		0.000	02:03
Diagnostics Check		OK	02:03

Data Download Date/Time: 01/30/2019 07:50

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLYNN,MICHAEL,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FOLADVANDI,M
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 12/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: DUNN

DATE/TIME OF TEST: 01/27/2019 02:52:16

Diagnostics Check		OK	02:45
Air Blank		0.000	02:45
Control Test		0.080	02:46
Air Blank		0.000	02:46
Subject Sample	# 1	0.262	02:47
Breath Volume		5.625	02:47
Air Blank		0.000	02:48
Air Blank		0.000	02:49
Subject Sample	# 2	0.257	02:50
Breath Volume		5.289	02:50
Air Blank		0.000	02:51
Control Test		0.078	02:51
Air Blank		0.000	02:52
Diagnostics Check		OK	02:52

Data Download Date/Time: 01/30/2019 07:50

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDEZ,DEPUTY,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MCMENAMIN,OFFICER
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 12/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:54

SUBJECT'S LAST NAME: MINGO

DATE/TIME OF TEST: 01/27/2019 19:22:45

Diagnostics Check		OK	19:14
Air Blank		0.000	19:15
Control Test		0.080	19:15
Air Blank		0.000	19:16
Subject Sample	# 1	0.293	19:17
Breath Volume		2.285	19:17
Air Blank		0.000	19:18
Air Blank		0.000	19:20
Subject Sample	# 2	0.301	19:21
Breath Volume		1.707	19:21
Air Blank		0.000	19:21
Control Test		0.078	19:22
Air Blank		0.000	19:22
Diagnostics Check		OK	19:22

Data Download Date/Time: 01/30/2019 07:50

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GELLER,DEBORAH,A
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: CARAVELLA,OFFICER
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 12/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 07:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:58

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 01/28/2019 09:47:09

Diagnostics Check		OK	09:40
Air Blank		0.000	09:40
Control Test		0.080	09:41
Air Blank		0.000	09:41
Subject Sample	# 1	0.000	09:42
Breath Volume		2.531	09:42
Air Blank		0.000	09:43
Air Blank		0.000	09:45
Subject Sample	# 2	0.000	09:45
Breath Volume		1.593	09:45
Air Blank		0.000	09:46
Control Test		0.080	09:46
Air Blank		0.000	09:47
Diagnostics Check		OK	09:47

Data Download Date/Time: 01/30/2019 07:50

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEDOREK,JEFFREY,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: ST JEAN,ERIC
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 12/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: LISITSIN

DATE/TIME OF TEST: 01/30/2019 03:07:56

Diagnostics Check		OK	03:01
Air Blank		0.000	03:01
Control Test		0.080	03:02
Air Blank		0.000	03:02
Subject Sample	# 1	0.090	03:02
Breath Volume		1.261	03:02
Air Blank		0.000	03:03
Air Blank		0.000	03:05
Subject Sample	# 2	0.092	03:06
Breath Volume		2.214	03:06
Air Blank		0.000	03:07
Control Test		0.079	03:07
Air Blank		0.000	03:07
Diagnostics Check		OK	03:07

Data Download Date/Time: 01/30/2019 07:50

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUMANN,DEP,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SCHUMANN,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: MOORMAN

DATE/TIME OF TEST: 01/30/2019 03:57:09

Diagnostics Check		OK	03:50
Air Blank		0.000	03:51
Control Test		0.080	03:51
Air Blank		0.000	03:52
Subject Sample	# 1	0.216	03:52
Breath Volume		2.019	03:52
Air Blank		0.000	03:53
Air Blank		0.000	03:55
Subject Sample	# 2	0.214	03:55
Breath Volume		2.304	03:55
Air Blank		0.000	03:56
Control Test		0.080	03:56
Air Blank		0.000	03:57
Diagnostics Check		OK	03:57

Data Download Date/Time: 01/30/2019 07:50

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORALES,RONNIE,
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: MORALES,RONNIE
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: CONKLIN

DATE/TIME OF TEST: 01/01/2019 01:19:32

Diagnostics Check		OK	01:12
Air Blank		0.000	01:13
Control Test		0.081	01:13
Air Blank		0.000	01:14
Subject Sample	# 1	0.000	01:14
Breath Volume		4.316	01:14
Air Blank		0.000	01:15
Air Blank		0.000	01:17
Subject Sample	# 2	0.000	01:18
Breath Volume		3.554	01:18
Air Blank		0.000	01:18
Control Test		0.080	01:18
Air Blank		0.000	01:19
Diagnostics Check		OK	01:19

Data Download Date/Time: 01/02/2019 15:25

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STROUP,STEVE,A
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: STROUP,STEVE
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 01/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 01/09/2019 02:06:33

Diagnostics Check	OK	02:05	
Air Blank		02:06	RFI Detect
Air Blank	0.000	02:06	

Data Download Date/Time: 02/04/2019 10:56

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STROUP,STEVE,A
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: STROUP,STEVE
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 01/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 01/09/2019 02:12:38

Diagnostics Check		OK	02:09	
Air Blank		0.000	02:10	
Control Test		0.080	02:10	
Air Blank		0.000	02:10	
Subject Sample	# 1		02:11	Subject Test Refused
Breath Volume		0.000	02:11	
Air Blank		0.000	02:11	
Control Test		0.079	02:12	
Air Blank		0.000	02:12	
Diagnostics Check		OK	02:12	

Data Download Date/Time: 02/04/2019 10:56

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORTES,VICTOR,E
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: ABDELMESSIH,ADAM
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 01/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 16:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:45

SUBJECT'S LAST NAME: REYNOLDS

DATE/TIME OF TEST: 01/09/2019 17:35:22

Diagnostics Check		OK	17:27
Air Blank		0.000	17:28
Control Test		0.080	17:28
Air Blank		0.000	17:28
Subject Sample	# 1	0.296	17:30
Breath Volume		1.523	17:30
Air Blank		0.000	17:31
Air Blank		0.000	17:33
Subject Sample	# 2	0.303	17:33
Breath Volume		1.472	17:33
Air Blank		0.000	17:34
Control Test		0.080	17:34
Air Blank		0.000	17:35
Diagnostics Check		OK	17:35

Data Download Date/Time: 02/04/2019 10:56

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIEVES,MATTHEW,J
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: STARLING,WILLIAM
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 01/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 01/13/2019 03:44:18

Diagnostics Check		OK	03:40	
Air Blank		0.000	03:41	
Control Test		0.081	03:41	
Air Blank		0.000	03:42	
Subject Sample	# 1		03:42	Subject Test Refused
Breath Volume		0.000	03:42	
Air Blank		0.000	03:43	
Control Test		0.081	03:43	
Air Blank		0.000	03:44	
Diagnostics Check		OK	03:44	

Data Download Date/Time: 02/04/2019 10:56

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HACKETT,SEAN,J
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: ALLRED,JAMES
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 01/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: ESCOBAR

DATE/TIME OF TEST: 01/24/2019 00:39:38

Diagnostics Check		OK	00:26	
Air Blank		0.000	00:27	
Control Test		0.080	00:27	
Air Blank		0.000	00:27	
Subject Sample	# 1	0.213	00:30	
Breath Volume		3.445	00:30	
Air Blank		0.000	00:31	
Air Blank		0.000	00:33	
Subject Sample	# 2	0.184	00:34	
Breath Volume		2.640	00:34	
Air Blank		0.000	00:35	
Air Blank		0.000	00:37	
Subject Sample	# 3	0.183	00:37	
Breath Volume		2.496	00:37	
Air Blank		0.000	00:38	
Control Test		0.074	00:39	Control Outside Tolerance
Air Blank		0.000	00:39	

Data Download Date/Time: 02/04/2019 10:56

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HACKETT,SEAN,J
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: ALLRED,JAMES
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 01/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: MICHAELS

DATE/TIME OF TEST: 01/30/2019 03:22:31

Diagnostics Check		OK	03:19	
Air Blank		0.000	03:19	
Control Test		0.081	03:20	
Air Blank		0.000	03:20	
Subject Sample	# 1		03:21	Subject Test Refused
Breath Volume		0.000	03:21	
Air Blank		0.000	03:21	
Control Test		0.081	03:21	
Air Blank		0.000	03:22	
Diagnostics Check		OK	03:22	

Data Download Date/Time: 02/04/2019 10:56

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUMANN,DEP,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SCHUMANN,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: QUINN

DATE/TIME OF TEST: 01/01/2019 02:55:50

Diagnostics Check		OK	02:48
Air Blank		0.000	02:49
Control Test		0.080	02:49
Air Blank		0.000	02:50
Subject Sample	# 1	0.183	02:50
Breath Volume		1.785	02:50
Air Blank		0.000	02:51
Air Blank		0.000	02:53
Subject Sample	# 2	0.169	02:54
Breath Volume		2.214	02:54
Air Blank		0.000	02:54
Control Test		0.078	02:55
Air Blank		0.000	02:55
Diagnostics Check		OK	02:55

Data Download Date/Time: 01/30/2019 09:05

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUMANN,DEP,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: CANTWELL,J
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:42

SUBJECT'S LAST NAME: ISSACS THOMAS

DATE/TIME OF TEST: 01/22/2019 20:13:09

Diagnostics Check		OK	20:06	
Air Blank		0.000	20:07	
Control Test		0.080	20:07	
Air Blank		0.000	20:07	
Subject Sample	# 1	0.135	20:09	
Breath Volume		3.070	20:09	
Air Blank		0.000	20:09	
Air Blank		0.000	20:11	
Subject Sample	# 2	0.134	20:12	
Breath Volume		2.621	20:12	
Air Blank			20:12	RFI Detect
Air Blank		0.000	20:13	

Data Download Date/Time: 01/30/2019 09:05

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUMANN,DEP,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SCHUMANN,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:08

SUBJECT'S LAST NAME: HERRON

DATE/TIME OF TEST: 01/23/2019 19:36:27

Diagnostics Check		OK	19:30
Air Blank		0.000	19:30
Control Test		0.080	19:31
Air Blank		0.000	19:31
Subject Sample	# 1	0.131	19:32
Breath Volume		1.363	19:32
Air Blank		0.000	19:32
Air Blank		0.000	19:34
Subject Sample	# 2	0.127	19:34
Breath Volume		1.609	19:34
Air Blank		0.000	19:35
Control Test		0.080	19:35
Air Blank		0.000	19:36
Diagnostics Check		OK	19:36

Data Download Date/Time: 01/30/2019 09:05

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUMANN,DEP,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: ROBERTS,PJ
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: SBERNA

DATE/TIME OF TEST: 01/26/2019 01:39:42

Diagnostics Check		OK	01:32
Air Blank		0.000	01:33
Control Test		0.081	01:33
Air Blank		0.000	01:33
Subject Sample	# 1	0.097	01:35
Breath Volume		1.761	01:35
Air Blank		0.000	01:35
Air Blank		0.000	01:37
Subject Sample	# 2	0.098	01:38
Breath Volume		1.730	01:38
Air Blank		0.000	01:38
Control Test		0.081	01:39
Air Blank		0.000	01:39
Diagnostics Check		OK	01:39

Data Download Date/Time: 01/30/2019 09:05

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSTYAL,MATT,A
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KOSTYAL,MATT
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: SAVAGE

DATE/TIME OF TEST: 01/26/2019 02:39:41

Diagnostics Check		OK	02:28	
Air Blank		0.000	02:28	
Control Test		0.081	02:28	
Air Blank		0.000	02:29	
Subject Sample	# 1		02:32	No Sample Provided
Breath Volume		0.000	02:32	
Air Blank		0.000	02:32	
Air Blank		0.000	02:34	
Subject Sample	# 2	0.064	02:35	
Breath Volume		2.027	02:35	
Air Blank		0.000	02:35	
Air Blank		0.000	02:37	
Subject Sample	# 3	0.060	02:38	
Breath Volume		1.523	02:38	
Air Blank		0.000	02:38	
Control Test		0.082	02:39	
Air Blank		0.000	02:39	
Diagnostics Check		OK	02:39	

Data Download Date/Time: 01/30/2019 08:45

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JATEFF,TIFFANY,J
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: JATEFF,TIFFANY
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 539031
DRY GAS EXPIRY DATE: 08/25/2019
VIOLATION CODE: DUI
ARREST TIME: 16:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:18

SUBJECT'S LAST NAME: CINTRON

DATE/TIME OF TEST: 01/20/2019 17:57:03

Diagnostics Check		OK	17:50
Air Blank		0.000	17:51
Control Test		0.082	17:51
Air Blank		0.000	17:51
Subject Sample	# 1	0.000	17:52
Breath Volume		1.664	17:52
Air Blank		0.000	17:53
Air Blank		0.000	17:55
Subject Sample	# 2	0.000	17:55
Breath Volume		2.546	17:55
Air Blank		0.000	17:56
Control Test		0.084	17:56
Air Blank		0.000	17:56
Diagnostics Check		OK	17:57

Data Download Date/Time: 01/23/2019 01:47

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: HALLAWAY

DATE/TIME OF TEST: 01/01/2019 00:58:31

Diagnostics Check		OK	00:52
Air Blank		0.000	00:52
Control Test		0.079	00:52
Air Blank		0.000	00:53
Subject Sample	# 1	0.133	00:54
Breath Volume		3.328	00:54
Air Blank		0.000	00:54
Air Blank		0.000	00:56
Subject Sample	# 2	0.125	00:57
Breath Volume		1.992	00:57
Air Blank		0.000	00:57
Control Test		0.079	00:57
Air Blank		0.000	00:58
Diagnostics Check		OK	00:58

Data Download Date/Time: 01/23/2019 21:39

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,KENDRICK,T
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: TURNER,KENDRICK
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: GLOVER

DATE/TIME OF TEST: 01/02/2019 03:35:46

Diagnostics Check		OK	03:28
Air Blank		0.000	03:29
Control Test		0.080	03:29
Air Blank		0.000	03:30
Subject Sample	# 1	0.225	03:30
Breath Volume		1.738	03:30
Air Blank		0.000	03:31
Air Blank		0.000	03:33
Subject Sample	# 2	0.217	03:34
Breath Volume		1.394	03:34
Air Blank		0.000	03:34
Control Test		0.075	03:35
Air Blank		0.000	03:35
Diagnostics Check		OK	03:35

Data Download Date/Time: 01/23/2019 21:39

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: CARPENTER

DATE/TIME OF TEST: 01/09/2019 23:50:09

Diagnostics Check		OK	23:44
Air Blank		0.000	23:44
Control Test		0.080	23:44
Air Blank		0.000	23:45
Subject Sample	# 1	0.324	23:45
Breath Volume		4.359	23:45
Air Blank		0.000	23:46
Air Blank		0.000	23:48
Subject Sample	# 2	0.317	23:48
Breath Volume		4.355	23:48
Air Blank		0.000	23:49
Control Test		0.079	23:49
Air Blank		0.000	23:50
Diagnostics Check		OK	23:50

Data Download Date/Time: 01/23/2019 21:39

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,CHRISTOPHER,
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GUYTON,JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: PIERCE

DATE/TIME OF TEST: 01/17/2019 03:20:08

Diagnostics Check		OK	03:15	
Air Blank		0.000	03:16	
Control Test		0.080	03:16	
Air Blank		0.000	03:17	
Subject Sample	# 1		03:18	Subject Test Refused
Breath Volume		0.000	03:18	
Air Blank		0.000	03:19	
Control Test		0.081	03:19	
Air Blank		0.000	03:20	
Diagnostics Check		OK	03:20	

Data Download Date/Time: 01/23/2019 21:39

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WENDT,HENDRIK,J
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: WENDT,HENDRIK
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: GREGORIE

DATE/TIME OF TEST: 01/19/2019 04:24:06

Diagnostics Check		OK	04:16
Air Blank		0.000	04:16
Control Test		0.080	04:16
Air Blank		0.000	04:17
Subject Sample	# 1	0.187	04:18
Breath Volume		2.074	04:18
Air Blank		0.000	04:19
Air Blank		0.000	04:21
Subject Sample	# 2	0.179	04:22
Breath Volume		1.609	04:22
Air Blank		0.000	04:23
Control Test		0.079	04:23
Air Blank		0.000	04:24
Diagnostics Check		OK	04:24

Data Download Date/Time: 01/23/2019 21:39

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,KENDRICK,T
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: TURNER,KENDRICK
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:12

SUBJECT'S LAST NAME: RIFFEE

DATE/TIME OF TEST: 01/21/2019 05:43:55

Diagnostics Check		OK	05:32	
Air Blank		0.000	05:33	
Control Test		0.081	05:33	
Air Blank		0.000	05:33	
Subject Sample	# 1		05:36	No Sample Provided
Breath Volume		0.000	05:36	
Air Blank		0.000	05:37	
Air Blank		0.000	05:39	
Subject Sample	# 2		05:42	No Sample Provided
Breath Volume		0.000	05:42	
Air Blank		0.000	05:43	
Control Test		0.081	05:43	
Air Blank		0.000	05:43	
Diagnostics Check		OK	05:43	

Data Download Date/Time: 01/23/2019 21:39

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,CHRISTOPHER,
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GUYTON,JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:24

SUBJECT'S LAST NAME: PATE

DATE/TIME OF TEST: 01/22/2019 21:58:00

Diagnostics Check		OK	21:45	
Air Blank		0.000	21:45	
Control Test		0.081	21:45	
Air Blank		0.000	21:46	
Subject Sample	# 1	0.026	21:49	Volume Not Met
Breath Volume		0.296	21:49	
Air Blank		0.000	21:50	
Air Blank		0.000	21:51	
Subject Sample	# 2	0.027	21:53	
Breath Volume		1.621	21:53	
Air Blank		0.000	21:54	
Air Blank		0.000	21:55	
Subject Sample	# 3	0.027	21:56	
Breath Volume		2.699	21:56	
Air Blank		0.000	21:57	
Control Test		0.080	21:57	
Air Blank		0.000	21:57	
Diagnostics Check		OK	21:58	

Data Download Date/Time: 01/23/2019 21:39

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRINSKO,THOMAS,K
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GOUGH,WILLIAM
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 01/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 08:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:55

SUBJECT'S LAST NAME: ROWDEN

DATE/TIME OF TEST: 01/24/2019 09:18:34

Diagnostics Check		OK	09:15	
Air Blank		0.000	09:16	
Control Test		0.079	09:16	
Air Blank		0.000	09:16	
Subject Sample	# 1		09:17	Subject Test Refused
Breath Volume		0.000	09:17	
Air Blank		0.000	09:17	
Control Test		0.079	09:17	
Air Blank		0.000	09:18	
Diagnostics Check		OK	09:18	

Data Download Date/Time: 02/21/2019 23:05

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,KENDRICK,T
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: TURNER,KENDRICK
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 01/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: MEDINA

DATE/TIME OF TEST: 01/25/2019 02:14:45

Diagnostics Check		OK	02:07
Air Blank		0.000	02:08
Control Test		0.080	02:08
Air Blank		0.000	02:09
Subject Sample	# 1	0.176	02:09
Breath Volume		3.457	02:09
Air Blank		0.000	02:10
Air Blank		0.000	02:12
Subject Sample	# 2	0.183	02:13
Breath Volume		1.222	02:13
Air Blank		0.000	02:13
Control Test		0.078	02:14
Air Blank		0.000	02:14
Diagnostics Check		OK	02:14

Data Download Date/Time: 02/21/2019 23:05

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,CHRISTOPHER,
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: TURNER,KENDRICK
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 01/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: PATTERSON

DATE/TIME OF TEST: 01/26/2019 01:27:15

Diagnostics Check		OK	01:20
Air Blank		0.000	01:20
Control Test		0.081	01:21
Air Blank		0.000	01:21
Subject Sample	# 1	0.165	01:22
Breath Volume		1.296	01:22
Air Blank		0.000	01:23
Air Blank		0.000	01:25
Subject Sample	# 2	0.163	01:25
Breath Volume		1.574	01:25
Air Blank		0.000	01:26
Control Test		0.080	01:26
Air Blank		0.000	01:27
Diagnostics Check		OK	01:27

Data Download Date/Time: 02/21/2019 23:05

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,CHRISTOPHER,
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GUYTON,JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 01/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 01/26/2019 23:21:28

Diagnostics Check		OK	23:16	
Air Blank		0.000	23:17	
Control Test		0.080	23:17	
Air Blank		0.000	23:17	
Subject Sample	# 1		23:20	Subject Test Refused
Breath Volume		0.000	23:20	
Air Blank		0.000	23:20	
Control Test		0.081	23:20	
Air Blank		0.000	23:21	
Diagnostics Check		OK	23:21	

Data Download Date/Time: 02/21/2019 23:05

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELCH,CHAD,J
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GUYTON,JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 01/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: HENDERSON

DATE/TIME OF TEST: 01/27/2019 03:26:26

Diagnostics Check		OK	03:15	
Air Blank		0.000	03:15	
Control Test		0.080	03:15	
Air Blank		0.000	03:16	
Subject Sample	# 1		03:19	No Sample Provided
Breath Volume		0.000	03:19	
Air Blank		0.000	03:19	
Air Blank		0.000	03:21	
Subject Sample	# 2		03:25	No Sample Provided
Breath Volume		0.000	03:25	
Air Blank		0.000	03:25	
Control Test		0.080	03:25	
Air Blank		0.000	03:26	
Diagnostics Check		OK	03:26	

Data Download Date/Time: 02/21/2019 23:05

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELCH,CHAD,J
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GUYTON,JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 01/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: WOOLDRIDGE JR

DATE/TIME OF TEST: 01/28/2019 01:50:39

Diagnostics Check		OK	01:44
Air Blank		0.000	01:44
Control Test		0.080	01:45
Air Blank		0.000	01:45
Subject Sample	# 1	0.195	01:46
Breath Volume		4.230	01:46
Air Blank		0.000	01:46
Air Blank		0.000	01:48
Subject Sample	# 2	0.189	01:49
Breath Volume		4.027	01:49
Air Blank		0.000	01:49
Control Test		0.079	01:50
Air Blank		0.000	01:50
Diagnostics Check		OK	01:50

Data Download Date/Time: 02/21/2019 23:05

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,KENDRICK,T
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: TURNER,KENDRICK
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 01/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: CHRISTENSEN

DATE/TIME OF TEST: 01/29/2019 02:26:08

Diagnostics Check		OK	02:14	
Air Blank		0.000	02:15	
Control Test		0.081	02:15	
Air Blank		0.000	02:16	
Subject Sample	# 1		02:19	No Sample Provided
Breath Volume		0.000	02:19	
Air Blank		0.000	02:19	
Air Blank		0.000	02:21	
Subject Sample	# 2		02:24	No Sample Provided
Breath Volume		0.000	02:24	
Air Blank		0.000	02:25	
Control Test		0.082	02:25	
Air Blank		0.000	02:26	
Diagnostics Check		OK	02:26	

Data Download Date/Time: 02/21/2019 23:05

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOWLING,DAKOTA,A
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: ROSA,CHANCE
ARRESTING OFFICER AGENCY: Williston PD
LAST AGENCY INSPECTION: 01/18/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: AG805701
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: DODD

DATE/TIME OF TEST: 01/28/2019 00:52:38

Diagnostics Check		OK	00:46
Air Blank		0.000	00:46
Control Test		0.080	00:46
Air Blank		0.000	00:47
Subject Sample	# 1	0.218	00:48
Breath Volume		3.578	00:48
Air Blank		0.000	00:48
Air Blank		0.000	00:50
Subject Sample	# 2	0.208	00:50
Breath Volume		1.980	00:50
Air Blank		0.000	00:51
Control Test		0.078	00:52
Air Blank		0.000	00:52
Diagnostics Check		OK	00:52

Data Download Date/Time: 02/10/2019 15:20

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINKOFER,Z,R
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: LONG,K
ARRESTING OFFICER AGENCY: Williston PD
LAST AGENCY INSPECTION: 01/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805701
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: LUNDEBY

DATE/TIME OF TEST: 01/29/2019 04:41:24

Diagnostics Check		OK	04:34
Air Blank		0.000	04:35
Control Test		0.080	04:35
Air Blank		0.000	04:36
Subject Sample	# 1	0.000	04:36
Breath Volume		1.734	04:36
Air Blank		0.000	04:37
Air Blank		0.000	04:39
Subject Sample	# 2	0.000	04:40
Breath Volume		1.855	04:40
Air Blank		0.000	04:40
Control Test		0.079	04:40
Air Blank		0.000	04:41
Diagnostics Check		OK	04:41

Data Download Date/Time: 02/10/2019 15:20

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINKOFER,ZACHARY,R
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: DRENNAN,D
ARRESTING OFFICER AGENCY: Williston PD
LAST AGENCY INSPECTION: 01/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805701
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 18:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:28

SUBJECT'S LAST NAME: DURAN

DATE/TIME OF TEST: 01/29/2019 19:53:05

Diagnostics Check		OK	19:50	
Air Blank		0.000	19:50	
Control Test		0.079	19:51	
Air Blank		0.000	19:51	
Subject Sample	# 1	0.000	19:52	
Breath Volume		1.535	19:52	
Air Blank			19:52	RFI Detect
Air Blank			19:52	RFI Detect
Air Blank		0.000	19:53	

Data Download Date/Time: 02/10/2019 15:20

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINKOFER,ZACHARY,R
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: DRENNAN,D
ARRESTING OFFICER AGENCY: Williston PD
LAST AGENCY INSPECTION: 01/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805701
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 18:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:28

SUBJECT'S LAST NAME: DURAN

DATE/TIME OF TEST: 01/29/2019 20:02:57

Diagnostics Check		OK	19:56
Air Blank		0.000	19:57
Control Test		0.079	19:57
Air Blank		0.000	19:57
Subject Sample	# 1	0.000	19:58
Breath Volume		1.175	19:58
Air Blank		0.000	19:58
Air Blank		0.000	20:00
Subject Sample	# 2	0.000	20:01
Breath Volume		1.789	20:01
Air Blank		0.000	20:02
Control Test		0.079	20:02
Air Blank		0.000	20:02
Diagnostics Check		OK	20:02

Data Download Date/Time: 02/10/2019 15:20

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICOTTILLI,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI,JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: ALEXANDER

DATE/TIME OF TEST: 01/01/2019 01:37:29

Diagnostics Check		OK	01:29
Air Blank		0.000	01:30
Control Test		0.081	01:30
Air Blank		0.000	01:31
Subject Sample	# 1	0.094	01:32
Breath Volume		1.726	01:32
Air Blank		0.000	01:33
Air Blank		0.000	01:35
Subject Sample	# 2	0.090	01:35
Breath Volume		1.492	01:35
Air Blank		0.000	01:36
Control Test		0.081	01:36
Air Blank		0.000	01:37
Diagnostics Check		OK	01:37

Data Download Date/Time: 01/30/2019 22:08

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICOTTILLI,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DEERMAN,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: BRITT

DATE/TIME OF TEST: 01/01/2019 02:09:58

Diagnostics Check		OK	02:07	
Air Blank		0.000	02:07	
Control Test		0.081	02:07	
Air Blank		0.000	02:08	
Subject Sample	# 1		02:08	Subject Test Refused
Breath Volume		0.000	02:08	
Air Blank		0.000	02:09	
Control Test		0.081	02:09	
Air Blank		0.000	02:09	
Diagnostics Check		OK	02:09	

Data Download Date/Time: 01/30/2019 22:08

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HANSEN,DEPUTY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 07:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:33

SUBJECT'S LAST NAME: PAAS

DATE/TIME OF TEST: 01/13/2019 08:00:15

Diagnostics Check		OK	07:53
Air Blank		0.000	07:54
Control Test		0.081	07:54
Air Blank		0.000	07:55
Subject Sample	# 1	0.233	07:55
Breath Volume		1.289	07:55
Air Blank		0.000	07:56
Air Blank		0.000	07:58
Subject Sample	# 2	0.215	07:58
Breath Volume		1.964	07:58
Air Blank		0.000	07:59
Control Test		0.081	07:59
Air Blank		0.000	08:00
Diagnostics Check		OK	08:00

Data Download Date/Time: 01/30/2019 22:08

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNAVELY,KENT,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SIEM,ALEX
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 11:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:28

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 01/17/2019 12:51:58

Diagnostics Check		OK	12:49	
Air Blank		0.000	12:49	
Control Test		0.082	12:49	
Air Blank		0.000	12:50	
Subject Sample	# 1		12:50	Subject Test Refused
Breath Volume		0.000	12:50	
Air Blank		0.000	12:51	
Control Test		0.082	12:51	
Air Blank		0.000	12:51	
Diagnostics Check		OK	12:51	

Data Download Date/Time: 01/30/2019 22:08

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVARTS,JASON,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: COOK,CHRISTOPHER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:32

SUBJECT'S LAST NAME: JAAP

DATE/TIME OF TEST: 01/22/2019 20:58:52

Diagnostics Check		OK	20:52
Air Blank		0.000	20:52
Control Test		0.081	20:53
Air Blank		0.000	20:53
Subject Sample	# 1	0.120	20:54
Breath Volume		1.472	20:54
Air Blank		0.000	20:54
Air Blank		0.000	20:56
Subject Sample	# 2	0.122	20:57
Breath Volume		1.554	20:57
Air Blank		0.000	20:57
Control Test		0.081	20:58
Air Blank		0.000	20:58
Diagnostics Check		OK	20:58

Data Download Date/Time: 01/30/2019 22:08

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLE,STEPHAN,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HOLE,STEPHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: LEGAGNEUR

DATE/TIME OF TEST: 01/26/2019 02:28:17

Diagnostics Check		OK	02:16	
Air Blank		0.000	02:17	
Control Test		0.082	02:17	
Air Blank		0.000	02:18	
Subject Sample	# 1		02:21	No Sample Provided
Breath Volume		0.000	02:21	
Air Blank		0.000	02:21	
Air Blank		0.000	02:23	
Subject Sample	# 2		02:26	No Sample Provided
Breath Volume		0.000	02:26	
Air Blank		0.000	02:27	
Control Test		0.082	02:27	
Air Blank		0.000	02:28	
Diagnostics Check		OK	02:28	

Data Download Date/Time: 01/30/2019 22:08

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

INSTRUMENT REGISTERED TO: NAS MAYPORT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OCASIO,JOSE,A
BREATH TEST OPERATOR AGENCY: Mayport Naval Base
ARRESTING OFFICER: JILES,D
ARRESTING OFFICER AGENCY: Mayport Naval Base
LAST AGENCY INSPECTION: 12/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 01/12/2019 04:05:08

Diagnostics Check		OK	03:58
Air Blank		0.000	03:58
Control Test		0.079	03:59
Air Blank		0.000	03:59
Subject Sample	# 1	0.110	04:00
Breath Volume		1.160	04:00
Air Blank		0.000	04:00
Air Blank		0.000	04:02
Subject Sample	# 2	0.107	04:03
Breath Volume		1.437	04:03
Air Blank		0.000	04:04
Control Test		0.079	04:04
Air Blank		0.000	04:05
Diagnostics Check		OK	04:05

Data Download Date/Time: 01/31/2019 14:48

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

INSTRUMENT REGISTERED TO: NAS MAYPORT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOOS,JAMES,P
BREATH TEST OPERATOR AGENCY: Mayport Naval Base
ARRESTING OFFICER: LOPEZ,RAYMOND
ARRESTING OFFICER AGENCY: Mayport Naval Base
LAST AGENCY INSPECTION: 12/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: ARDOIN

DATE/TIME OF TEST: 01/18/2019 03:34:46

Diagnostics Check		OK	03:27	
Air Blank		0.000	03:28	
Control Test		0.080	03:28	
Air Blank		0.000	03:28	
Subject Sample	# 1	0.127	03:30	
Breath Volume		3.265	03:30	
Air Blank		0.000	03:30	
Air Blank		0.000	03:32	
Subject Sample	# 2	0.129	03:33	
Breath Volume		3.664	03:33	
Air Blank		0.000	03:33	
Control Test		0.079	03:34	RFI Detect
Air Blank		0.000	03:34	

Data Download Date/Time: 01/31/2019 14:48

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

INSTRUMENT REGISTERED TO: NAS MAYPORT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OCASIO,JOSE,A
BREATH TEST OPERATOR AGENCY: Mayport Naval Base
ARRESTING OFFICER: TUER,DAVID
ARRESTING OFFICER AGENCY: Mayport Naval Base
LAST AGENCY INSPECTION: 12/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: AYALA

DATE/TIME OF TEST: 01/26/2019 04:00:42

Diagnostics Check		OK	03:57	
Air Blank		0.000	03:58	
Control Test		0.080	03:58	
Air Blank		0.000	03:58	
Subject Sample	# 1		03:59	Subject Test Refused
Breath Volume		0.000	03:59	
Air Blank		0.000	03:59	
Control Test		0.080	04:00	
Air Blank		0.000	04:00	
Diagnostics Check		OK	04:00	

Data Download Date/Time: 01/31/2019 14:48

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

INSTRUMENT REGISTERED TO: NAS MAYPORT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OCASIO,JOSE,A
BREATH TEST OPERATOR AGENCY: Mayport Naval Base
ARRESTING OFFICER: TUER,DAVID
ARRESTING OFFICER AGENCY: Mayport Naval Base
LAST AGENCY INSPECTION: 12/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: AYALA

DATE/TIME OF TEST: 01/26/2019 04:26:10

Diagnostics Check	OK	04:25	
Air Blank	0.000	04:25	
Control Test	0.000	04:25	RFI Detect
Air Blank	0.000	04:26	

Data Download Date/Time: 01/31/2019 14:48

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ROCCHILD,MIGUEL
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: RUIZLOPEZ

DATE/TIME OF TEST: 01/01/2019 00:00:53

Diagnostics Check		OK	23:54
Air Blank		0.000	23:55
Control Test		0.082	23:55
Air Blank		0.000	23:56
Subject Sample	# 1	0.205	23:56
Breath Volume		1.867	23:56
Air Blank		0.000	23:57
Air Blank		0.000	23:58
Subject Sample	# 2	0.209	23:59
Breath Volume		1.859	23:59
Air Blank		0.000	23:59
Control Test		0.081	00:00
Air Blank		0.000	00:00
Diagnostics Check		OK	00:00

Data Download Date/Time: 01/03/2019 13:47

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: REMBISLICARI,JASMINE
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: TOMASVICENTE

DATE/TIME OF TEST: 01/01/2019 03:27:42

Diagnostics Check		OK	03:20
Air Blank		0.000	03:20
Control Test		0.082	03:21
Air Blank		0.000	03:21
Subject Sample	# 1	0.149	03:22
Breath Volume		1.187	03:22
Air Blank		0.000	03:23
Air Blank		0.000	03:25
Subject Sample	# 2	0.153	03:26
Breath Volume		1.460	03:26
Air Blank		0.000	03:26
Control Test		0.081	03:27
Air Blank		0.000	03:27
Diagnostics Check		OK	03:27

Data Download Date/Time: 01/03/2019 13:47

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CORTES,ADAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:43

SUBJECT'S LAST NAME: SANCHEZ O

DATE/TIME OF TEST: 01/02/2019 22:13:06

Diagnostics Check		OK	22:03
Air Blank		0.000	22:04
Control Test		0.082	22:04
Air Blank		0.000	22:05
Subject Sample	# 1	0.327	22:05
Breath Volume		2.406	22:05
Air Blank		0.000	22:06
Air Blank		0.000	22:08
Subject Sample	# 2	0.359	22:08
Breath Volume		1.210	22:08
Air Blank		0.000	22:09
Air Blank		0.000	22:11
Subject Sample	# 3	0.358	22:11
Breath Volume		2.957	22:11
Air Blank		0.000	22:12
Control Test		0.080	22:12
Air Blank		0.000	22:13
Diagnostics Check		OK	22:13

Data Download Date/Time: 01/03/2019 13:47

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: FERRIS,JEFFREY
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: MICHEL

DATE/TIME OF TEST: 01/04/2019 03:49:39

Diagnostics Check		OK	03:43
Air Blank		0.000	03:44
Control Test		0.082	03:44
Air Blank		0.000	03:45
Subject Sample	# 1	0.000	03:45
Breath Volume		1.488	03:45
Air Blank		0.000	03:45
Air Blank		0.000	03:47
Subject Sample	# 2	0.000	03:48
Breath Volume		1.285	03:48
Air Blank		0.000	03:48
Control Test		0.082	03:49
Air Blank		0.000	03:49
Diagnostics Check		OK	03:49

Data Download Date/Time: 01/16/2019 11:46

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DAVILA,S
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 01/05/2019 00:56:04

Diagnostics Check		OK	00:44	
Air Blank		0.000	00:44	
Control Test		0.081	00:44	
Air Blank		0.000	00:45	
Subject Sample	# 1	0.135	00:45	
Breath Volume		1.722	00:45	
Air Blank		0.000	00:46	
Air Blank		0.000	00:48	
Subject Sample	# 2	0.137	00:51	Volume Not Met
Breath Volume		0.039	00:51	
Air Blank		0.000	00:52	
Air Blank		0.000	00:54	
Subject Sample	# 3	0.133	00:54	
Breath Volume		1.953	00:54	
Air Blank		0.000	00:55	
Control Test		0.079	00:55	
Air Blank		0.000	00:55	
Diagnostics Check		OK	00:56	

Data Download Date/Time: 01/16/2019 11:46

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CORSO,ANTHONY
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: FUENTES

DATE/TIME OF TEST: 01/05/2019 03:10:06

Diagnostics Check		OK	03:02
Air Blank		0.000	03:03
Control Test		0.081	03:03
Air Blank		0.000	03:04
Subject Sample	# 1	0.192	03:04
Breath Volume		3.859	03:04
Air Blank		0.000	03:05
Air Blank		0.000	03:07
Subject Sample	# 2	0.179	03:08
Breath Volume		2.828	03:08
Air Blank		0.000	03:09
Control Test		0.081	03:09
Air Blank		0.000	03:10
Diagnostics Check		OK	03:10

Data Download Date/Time: 01/16/2019 11:46

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PRICE,FORREST
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: MCGUIRK

DATE/TIME OF TEST: 01/06/2019 02:42:24

Diagnostics Check		OK	02:39	
Air Blank		0.000	02:40	
Control Test		0.082	02:40	
Air Blank		0.000	02:40	
Subject Sample	# 1		02:41	Subject Test Refused
Breath Volume		0.000	02:41	
Air Blank		0.000	02:41	
Control Test		0.082	02:41	
Air Blank		0.000	02:42	
Diagnostics Check		OK	02:42	

Data Download Date/Time: 01/16/2019 11:46

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KRALICH,BRENDEN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: HERRSCHAFT

DATE/TIME OF TEST: 01/06/2019 03:35:04

Diagnostics Check		OK	03:28
Air Blank		0.000	03:28
Control Test		0.082	03:28
Air Blank		0.000	03:29
Subject Sample	# 1	0.178	03:30
Breath Volume		1.230	03:30
Air Blank		0.000	03:30
Air Blank		0.000	03:32
Subject Sample	# 2	0.179	03:33
Breath Volume		2.445	03:33
Air Blank		0.000	03:34
Control Test		0.081	03:34
Air Blank		0.000	03:34
Diagnostics Check		OK	03:35

Data Download Date/Time: 01/16/2019 11:46

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: STANLEY,JONATHAN
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: ACOSTA

DATE/TIME OF TEST: 01/06/2019 05:06:32

Diagnostics Check		OK	04:57
Air Blank		0.000	04:57
Control Test		0.082	04:57
Air Blank		0.000	04:58
Subject Sample	# 1	0.138	05:01
Breath Volume		1.433	05:01
Air Blank		0.000	05:02
Air Blank		0.000	05:03
Subject Sample	# 2	0.135	05:04
Breath Volume		1.757	05:04
Air Blank		0.000	05:05
Control Test		0.080	05:05
Air Blank		0.000	05:06
Diagnostics Check		OK	05:06

Data Download Date/Time: 01/16/2019 11:46

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BINET,MARLON
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: PEREZ BAUTISTA

DATE/TIME OF TEST: 01/07/2019 00:20:32

Diagnostics Check		OK	00:14
Air Blank		0.000	00:14
Control Test		0.082	00:15
Air Blank		0.000	00:15
Subject Sample	# 1	0.227	00:16
Breath Volume		2.046	00:16
Air Blank		0.000	00:16
Air Blank		0.000	00:18
Subject Sample	# 2	0.230	00:19
Breath Volume		2.085	00:19
Air Blank		0.000	00:19
Control Test		0.081	00:19
Air Blank		0.000	00:20
Diagnostics Check		OK	00:20

Data Download Date/Time: 01/16/2019 11:46

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: FIGUEROA,HECTOR
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 01/07/2019 03:12:01

Diagnostics Check		OK	03:05
Air Blank		0.000	03:05
Control Test		0.082	03:06
Air Blank		0.000	03:06
Subject Sample	# 1	0.179	03:07
Breath Volume		2.824	03:07
Air Blank		0.000	03:07
Air Blank		0.000	03:09
Subject Sample	# 2	0.164	03:10
Breath Volume		1.410	03:10
Air Blank		0.000	03:11
Control Test		0.081	03:11
Air Blank		0.000	03:11
Diagnostics Check		OK	03:12

Data Download Date/Time: 01/16/2019 11:46

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: NUNEZ,ANDREW
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: DUENASFORERO

DATE/TIME OF TEST: 01/07/2019 03:38:07

Diagnostics Check		OK	03:31
Air Blank		0.000	03:32
Control Test		0.080	03:32
Air Blank		0.000	03:32
Subject Sample	# 1	0.163	03:33
Breath Volume		3.523	03:33
Air Blank		0.000	03:34
Air Blank		0.000	03:35
Subject Sample	# 2	0.162	03:36
Breath Volume		3.585	03:36
Air Blank		0.000	03:37
Control Test		0.081	03:37
Air Blank		0.000	03:38
Diagnostics Check		OK	03:38

Data Download Date/Time: 01/16/2019 11:46

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LACENTRA,DONALD
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 06:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:12

SUBJECT'S LAST NAME: VANASDALE

DATE/TIME OF TEST: 01/07/2019 07:41:58

Diagnostics Check		OK	07:35
Air Blank		0.000	07:35
Control Test		0.082	07:35
Air Blank		0.000	07:36
Subject Sample	# 1	0.115	07:37
Breath Volume		2.175	07:37
Air Blank		0.000	07:38
Air Blank		0.000	07:40
Subject Sample	# 2	0.104	07:40
Breath Volume		1.378	07:40
Air Blank		0.000	07:41
Control Test		0.081	07:41
Air Blank		0.000	07:41
Diagnostics Check		OK	07:41

Data Download Date/Time: 01/16/2019 11:46

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AZEVEDO,FABIO
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:14

SUBJECT'S LAST NAME: MCLEAN

DATE/TIME OF TEST: 01/07/2019 19:50:34

Diagnostics Check		OK	19:39	
Air Blank		0.000	19:39	
Control Test		0.082	19:39	
Air Blank		0.000	19:40	
Subject Sample	# 1		19:43	No Sample Provided
Breath Volume		0.000	19:43	
Air Blank		0.000	19:44	
Air Blank		0.000	19:46	
Subject Sample	# 2		19:49	No Sample Provided
Breath Volume		0.000	19:49	
Air Blank		0.000	19:49	
Control Test		0.081	19:49	
Air Blank		0.000	19:50	
Diagnostics Check		OK	19:50	

Data Download Date/Time: 01/16/2019 11:46

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CILENTI,ANTHONY
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: DOOLEY

DATE/TIME OF TEST: 01/08/2019 02:59:57

Diagnostics Check		OK	02:52
Air Blank		0.000	02:52
Control Test		0.082	02:53
Air Blank		0.000	02:53
Subject Sample	# 1	0.111	02:55
Breath Volume		1.152	02:55
Air Blank		0.000	02:55
Air Blank		0.000	02:57
Subject Sample	# 2	0.113	02:58
Breath Volume		1.617	02:58
Air Blank		0.000	02:59
Control Test		0.081	02:59
Air Blank		0.000	02:59
Diagnostics Check		OK	02:59

Data Download Date/Time: 01/16/2019 11:46

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: NELSON JR,GREGG
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: CARBAJAL

DATE/TIME OF TEST: 01/08/2019 04:19:16

Diagnostics Check		OK	04:11
Air Blank		0.000	04:12
Control Test		0.082	04:12
Air Blank		0.000	04:13
Subject Sample	# 1	0.058	04:14
Breath Volume		2.285	04:14
Air Blank		0.000	04:14
Air Blank		0.000	04:16
Subject Sample	# 2	0.057	04:17
Breath Volume		2.019	04:17
Air Blank		0.000	04:18
Control Test		0.082	04:18
Air Blank		0.000	04:19
Diagnostics Check		OK	04:19

Data Download Date/Time: 01/16/2019 11:46

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GONCALVES,G
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: NGUYEN

DATE/TIME OF TEST: 01/10/2019 20:47:45

Diagnostics Check		OK	20:41
Air Blank		0.000	20:42
Control Test		0.082	20:42
Air Blank		0.000	20:43
Subject Sample	# 1	0.205	20:43
Breath Volume		3.257	20:43
Air Blank		0.000	20:44
Air Blank		0.000	20:45
Subject Sample	# 2	0.206	20:46
Breath Volume		3.675	20:46
Air Blank		0.000	20:46
Control Test		0.081	20:47
Air Blank		0.000	20:47
Diagnostics Check		OK	20:47

Data Download Date/Time: 01/16/2019 11:46

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CORSO,ANTHONY
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: AYALA

DATE/TIME OF TEST: 01/11/2019 05:05:38

Diagnostics Check		OK	04:58
Air Blank		0.000	04:59
Control Test		0.082	04:59
Air Blank		0.000	05:00
Subject Sample	# 1	0.078	05:01
Breath Volume		3.406	05:01
Air Blank		0.000	05:01
Air Blank		0.000	05:03
Subject Sample	# 2	0.075	05:04
Breath Volume		2.410	05:04
Air Blank		0.000	05:04
Control Test		0.081	05:05
Air Blank		0.000	05:05
Diagnostics Check		OK	05:05

Data Download Date/Time: 01/16/2019 11:46

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MOONDA,SHUAIB
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: REILLO

DATE/TIME OF TEST: 01/12/2019 04:20:07

Diagnostics Check		OK	04:14
Air Blank		0.000	04:14
Control Test		0.083	04:14
Air Blank		0.000	04:15
Subject Sample	# 1	0.118	04:15
Breath Volume		1.898	04:15
Air Blank		0.000	04:16
Air Blank		0.000	04:18
Subject Sample	# 2	0.117	04:18
Breath Volume		2.132	04:18
Air Blank		0.000	04:19
Control Test		0.081	04:19
Air Blank		0.000	04:20
Diagnostics Check		OK	04:20

Data Download Date/Time: 01/16/2019 11:46

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KNIGHT,RYAN
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: CORTESFIGUEROA

DATE/TIME OF TEST: 01/13/2019 04:29:27

Diagnostics Check		OK	04:22
Air Blank		0.000	04:23
Control Test		0.082	04:23
Air Blank		0.000	04:24
Subject Sample	# 1	0.140	04:25
Breath Volume		3.421	04:25
Air Blank		0.000	04:25
Air Blank		0.000	04:27
Subject Sample	# 2	0.135	04:27
Breath Volume		3.570	04:27
Air Blank		0.000	04:28
Control Test		0.080	04:28
Air Blank		0.000	04:29
Diagnostics Check		OK	04:29

Data Download Date/Time: 01/16/2019 11:46

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: REYNANTE,PORCHIA
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 11:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: ZITO

DATE/TIME OF TEST: 01/14/2019 00:23:10

Diagnostics Check	OK	00:21	
Air Blank	0.000	00:22	
Control Test	0.082	00:22	
Air Blank		00:22	RFI Detect
Air Blank	0.000	00:23	

Data Download Date/Time: 01/16/2019 11:46

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: REYNANTE,PORCHIA
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 11:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: ZITO

DATE/TIME OF TEST: 01/14/2019 00:32:17

Diagnostics Check		OK	00:24
Air Blank		0.000	00:25
Control Test		0.082	00:25
Air Blank		0.000	00:26
Subject Sample	# 1	0.078	00:27
Breath Volume		3.679	00:27
Air Blank		0.000	00:28
Air Blank		0.000	00:30
Subject Sample	# 2	0.081	00:30
Breath Volume		2.542	00:30
Air Blank		0.000	00:31
Control Test		0.082	00:31
Air Blank		0.000	00:32
Diagnostics Check		OK	00:32

Data Download Date/Time: 01/16/2019 11:46

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HERNANDEZ,YADRIEL
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: REDINGER

DATE/TIME OF TEST: 01/14/2019 05:26:40

Diagnostics Check		OK	05:14	
Air Blank		0.000	05:15	
Control Test		0.081	05:15	
Air Blank		0.000	05:16	
Subject Sample	# 1	0.199	05:19	Volume Not Met
Breath Volume		0.121	05:19	
Air Blank		0.000	05:20	
Air Blank		0.000	05:21	
Subject Sample	# 2	0.210	05:24	Volume Not Met
Breath Volume		0.589	05:24	
Air Blank		0.000	05:25	
Control Test		0.080	05:26	
Air Blank		0.000	05:26	
Diagnostics Check		OK	05:26	

Data Download Date/Time: 01/16/2019 11:46

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HERNANDEZ,YADRIEL
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: REDINGER

DATE/TIME OF TEST: 01/14/2019 05:36:22

Diagnostics Check		OK	05:27	
Air Blank		0.000	05:28	
Control Test		0.080	05:28	
Air Blank		0.000	05:29	
Subject Sample	# 1	0.202	05:32	Volume Not Met
Breath Volume		0.582	05:32	
Air Blank		0.000	05:33	
Air Blank		0.000	05:34	
Subject Sample	# 2		05:35	Subject Test Refused
Breath Volume		0.000	05:35	
Air Blank		0.000	05:35	
Control Test		0.079	05:35	
Air Blank		0.000	05:36	
Diagnostics Check		OK	05:36	

Data Download Date/Time: 01/16/2019 11:46

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LOPEZ,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: PICKETT

DATE/TIME OF TEST: 01/14/2019 22:33:31

Diagnostics Check		OK	22:27
Air Blank		0.000	22:27
Control Test		0.081	22:28
Air Blank		0.000	22:28
Subject Sample	# 1	0.151	22:29
Breath Volume		3.835	22:29
Air Blank		0.000	22:29
Air Blank		0.000	22:31
Subject Sample	# 2	0.148	22:31
Breath Volume		3.722	22:31
Air Blank		0.000	22:32
Control Test		0.082	22:32
Air Blank		0.000	22:33
Diagnostics Check		OK	22:33

Data Download Date/Time: 01/16/2019 11:46

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BEDOYA,HERNANDO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 01/15/2019 01:18:17

Diagnostics Check		OK	01:12
Air Blank		0.000	01:12
Control Test		0.082	01:13
Air Blank		0.000	01:13
Subject Sample	# 1	0.161	01:13
Breath Volume		3.554	01:13
Air Blank		0.000	01:14
Air Blank		0.000	01:16
Subject Sample	# 2	0.152	01:16
Breath Volume		1.796	01:16
Air Blank		0.000	01:17
Control Test		0.081	01:17
Air Blank		0.000	01:18
Diagnostics Check		OK	01:18

Data Download Date/Time: 01/16/2019 11:46

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KERN,ETHAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: BLAIR

DATE/TIME OF TEST: 01/15/2019 02:29:07

Diagnostics Check		OK	02:23
Air Blank		0.000	02:23
Control Test		0.083	02:23
Air Blank		0.000	02:24
Subject Sample	# 1	0.206	02:24
Breath Volume		3.300	02:24
Air Blank		0.000	02:25
Air Blank		0.000	02:27
Subject Sample	# 2	0.191	02:27
Breath Volume		1.730	02:27
Air Blank		0.000	02:28
Control Test		0.081	02:28
Air Blank		0.000	02:29
Diagnostics Check		OK	02:29

Data Download Date/Time: 01/16/2019 11:46

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DAVILA,SAMUEL
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: WATERHOUSE

DATE/TIME OF TEST: 01/16/2019 04:25:16

Diagnostics Check		OK	04:16
Air Blank		0.000	04:17
Control Test		0.082	04:17
Air Blank		0.000	04:18
Subject Sample	# 1	0.160	04:20
Breath Volume		1.582	04:20
Air Blank		0.000	04:21
Air Blank		0.000	04:23
Subject Sample	# 2	0.162	04:23
Breath Volume		3.679	04:23
Air Blank		0.000	04:24
Control Test		0.080	04:24
Air Blank		0.000	04:25
Diagnostics Check		OK	04:25

Data Download Date/Time: 01/16/2019 11:46

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ODEA,PATRICIA
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: HILLERY

DATE/TIME OF TEST: 01/17/2019 01:12:59

Diagnostics Check		OK	01:05
Air Blank		0.000	01:06
Control Test		0.082	01:06
Air Blank		0.000	01:07
Subject Sample	# 1	0.202	01:07
Breath Volume		2.726	01:07
Air Blank		0.000	01:08
Air Blank		0.000	01:10
Subject Sample	# 2	0.183	01:11
Breath Volume		1.187	01:11
Air Blank		0.000	01:12
Control Test		0.081	01:12
Air Blank		0.000	01:12
Diagnostics Check		OK	01:12

Data Download Date/Time: 01/24/2019 16:31

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KISSANE,BRIAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 01/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 01/17/2019 18:38:10

Diagnostics Check		OK	18:21	
Air Blank		0.000	18:22	
Control Test		0.082	18:22	
Air Blank		0.000	18:23	
Subject Sample	# 1	0.175	18:26	Volume Not Met
Breath Volume		0.386	18:26	
Air Blank		0.000	18:26	
Air Blank		0.000	18:28	
Subject Sample	# 2	0.181	18:30	
Breath Volume		1.742	18:30	
Air Blank		0.000	18:31	
Air Blank		0.000	18:33	
Subject Sample	# 3	0.162	18:36	Volume Not Met
Breath Volume		0.414	18:36	
Air Blank		0.000	18:37	
Control Test		0.080	18:37	
Air Blank		0.000	18:38	
Diagnostics Check		OK	18:38	

Data Download Date/Time: 01/24/2019 16:31

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KAREEM,LAITH
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 01/18/2019 22:03:10

Diagnostics Check		OK	21:56
Air Blank		0.000	21:57
Control Test		0.082	21:57
Air Blank		0.000	21:58
Subject Sample	# 1	0.173	21:58
Breath Volume		1.765	21:58
Air Blank		0.000	21:58
Air Blank		0.000	22:00
Subject Sample	# 2	0.166	22:01
Breath Volume		2.027	22:01
Air Blank		0.000	22:02
Control Test		0.082	22:02
Air Blank		0.000	22:03
Diagnostics Check		OK	22:03

Data Download Date/Time: 01/24/2019 16:31

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CALISTE,VIXON
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 01/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: HARDIN

DATE/TIME OF TEST: 01/19/2019 00:46:05

Diagnostics Check		OK	00:40
Air Blank		0.000	00:40
Control Test		0.082	00:40
Air Blank		0.000	00:41
Subject Sample	# 1	0.000	00:41
Breath Volume		2.500	00:41
Air Blank		0.000	00:42
Air Blank		0.000	00:44
Subject Sample	# 2	0.000	00:44
Breath Volume		2.578	00:44
Air Blank		0.000	00:45
Control Test		0.082	00:45
Air Blank		0.000	00:46
Diagnostics Check		OK	00:46

Data Download Date/Time: 01/24/2019 16:32

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BINET,MARLON
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 01/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 01/19/2019 02:11:39

Diagnostics Check		OK	02:03	
Air Blank		0.000	02:03	
Control Test		0.081	02:04	
Air Blank		0.000	02:04	
Subject Sample	# 1		02:07	No Sample Provided
Breath Volume		0.000	02:07	
Air Blank		0.000	02:08	
Air Blank		0.000	02:10	
Subject Sample	# 2		02:10	Subject Test Refused
Breath Volume		0.000	02:10	
Air Blank		0.000	02:10	
Control Test		0.081	02:11	
Air Blank		0.000	02:11	
Diagnostics Check		OK	02:11	

Data Download Date/Time: 01/24/2019 16:32

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PRICE,FORREST
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 01/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: HECTOR

DATE/TIME OF TEST: 01/19/2019 03:37:59

Diagnostics Check		OK	03:30
Air Blank		0.000	03:31
Control Test		0.081	03:31
Air Blank		0.000	03:31
Subject Sample	# 1	0.203	03:33
Breath Volume		1.738	03:33
Air Blank		0.000	03:33
Air Blank		0.000	03:35
Subject Sample	# 2	0.188	03:36
Breath Volume		1.132	03:36
Air Blank		0.000	03:37
Control Test		0.080	03:37
Air Blank		0.000	03:37
Diagnostics Check		OK	03:37

Data Download Date/Time: 01/24/2019 16:32

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CONWAY,C
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 01/19/2019 20:44:49

Diagnostics Check		OK	20:42	
Air Blank		0.000	20:43	
Control Test		0.081	20:43	
Air Blank		0.000	20:43	
Subject Sample	# 1		20:44	RFI Detect
Breath Volume		0.000	20:44	
Air Blank			20:44	RFI Detect
Air Blank			20:44	RFI Detect

Data Download Date/Time: 01/24/2019 16:32

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAWKINS,DEB
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: BENOIT

DATE/TIME OF TEST: 01/20/2019 00:40:46

Diagnostics Check		OK	00:34
Air Blank		0.000	00:35
Control Test		0.081	00:35
Air Blank		0.000	00:35
Subject Sample	# 1	0.156	00:36
Breath Volume		2.289	00:36
Air Blank		0.000	00:36
Air Blank		0.000	00:38
Subject Sample	# 2	0.158	00:39
Breath Volume		2.304	00:39
Air Blank		0.000	00:39
Control Test		0.082	00:40
Air Blank		0.000	00:40
Diagnostics Check		OK	00:40

Data Download Date/Time: 01/24/2019 16:32

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SOSA,EMMANUEL
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 01/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: SANTANA RAMIREZ

DATE/TIME OF TEST: 01/20/2019 01:21:31

Diagnostics Check		OK	01:15
Air Blank		0.000	01:15
Control Test		0.081	01:15
Air Blank		0.000	01:16
Subject Sample	# 1	0.063	01:17
Breath Volume		3.468	01:17
Air Blank		0.000	01:17
Air Blank		0.000	01:19
Subject Sample	# 2	0.061	01:20
Breath Volume		3.144	01:20
Air Blank		0.000	01:20
Control Test		0.081	01:20
Air Blank		0.000	01:21
Diagnostics Check		OK	01:21

Data Download Date/Time: 01/24/2019 16:32

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ROCCHILD,MIGUEL
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 01/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: PAREDES

DATE/TIME OF TEST: 01/20/2019 03:37:32

Diagnostics Check		OK	03:30
Air Blank		0.000	03:31
Control Test		0.081	03:31
Air Blank		0.000	03:32
Subject Sample	# 1	0.152	03:32
Breath Volume		1.339	03:32
Air Blank		0.000	03:32
Air Blank		0.000	03:34
Subject Sample	# 2	0.163	03:36
Breath Volume		1.273	03:36
Air Blank		0.000	03:36
Control Test		0.082	03:36
Air Blank		0.000	03:37
Diagnostics Check		OK	03:37

Data Download Date/Time: 01/24/2019 16:32

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BRYAN,ADAM
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 01/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 01/20/2019 05:41:54

Diagnostics Check		OK	05:36
Air Blank		0.000	05:36
Control Test		0.081	05:36
Air Blank		0.000	05:37
Subject Sample	# 1	0.178	05:37
Breath Volume		3.011	05:37
Air Blank		0.000	05:38
Air Blank		0.000	05:40
Subject Sample	# 2	0.181	05:40
Breath Volume		3.011	05:40
Air Blank		0.000	05:41
Control Test		0.080	05:41
Air Blank		0.000	05:41
Diagnostics Check		OK	05:41

Data Download Date/Time: 01/24/2019 16:32

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DAVILA,SAMUEL
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 01/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: JASTROWSKI

DATE/TIME OF TEST: 01/20/2019 19:32:03

Diagnostics Check		OK	19:26
Air Blank		0.000	19:26
Control Test		0.082	19:26
Air Blank		0.000	19:27
Subject Sample	# 1	0.236	19:27
Breath Volume		1.179	19:27
Air Blank		0.000	19:28
Air Blank		0.000	19:30
Subject Sample	# 2	0.231	19:30
Breath Volume		3.566	19:30
Air Blank		0.000	19:31
Control Test		0.081	19:31
Air Blank		0.000	19:31
Diagnostics Check		OK	19:32

Data Download Date/Time: 01/24/2019 16:32

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LEE,MIKEL
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 01/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: HATHCOTE

DATE/TIME OF TEST: 01/21/2019 22:47:27

Diagnostics Check		OK	22:41
Air Blank		0.000	22:41
Control Test		0.081	22:41
Air Blank		0.000	22:42
Subject Sample	# 1	0.245	22:43
Breath Volume		2.109	22:43
Air Blank		0.000	22:43
Air Blank		0.000	22:45
Subject Sample	# 2	0.261	22:45
Breath Volume		3.195	22:45
Air Blank		0.000	22:46
Control Test		0.083	22:46
Air Blank		0.000	22:47
Diagnostics Check		OK	22:47

Data Download Date/Time: 01/24/2019 16:32

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TOLIVER,JOVAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 01/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: VALLIER

DATE/TIME OF TEST: 01/21/2019 23:28:01

Diagnostics Check		OK	23:16	
Air Blank		0.000	23:16	
Control Test		0.082	23:17	
Air Blank		0.000	23:17	
Subject Sample	# 1	0.280	23:20	Volume Not Met
Breath Volume		0.179	23:20	
Air Blank		0.000	23:21	
Air Blank		0.000	23:23	
Subject Sample	# 2	0.270	23:26	Volume Not Met
Breath Volume		0.160	23:26	
Air Blank		0.000	23:27	
Control Test		0.081	23:27	
Air Blank		0.000	23:27	
Diagnostics Check		OK	23:28	

Data Download Date/Time: 01/24/2019 16:32

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VENTURA,RONNIE
ARRESTING OFFICER AGENCY: Edgewood PD
LAST AGENCY INSPECTION: 01/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 01/22/2019 01:57:03

Diagnostics Check		OK	01:54	
Air Blank		0.000	01:54	
Control Test		0.082	01:55	
Air Blank		0.000	01:55	
Subject Sample	# 1		01:55	Subject Test Refused
Breath Volume		0.000	01:55	
Air Blank		0.000	01:56	
Control Test		0.082	01:56	
Air Blank		0.000	01:56	
Diagnostics Check		OK	01:57	

Data Download Date/Time: 01/24/2019 16:32

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LAL,RAJIV
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 01/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: LARSON

DATE/TIME OF TEST: 01/22/2019 03:40:10

Diagnostics Check		OK	03:33
Air Blank		0.000	03:34
Control Test		0.082	03:34
Air Blank		0.000	03:34
Subject Sample	# 1	0.226	03:35
Breath Volume		2.113	03:35
Air Blank		0.000	03:36
Air Blank		0.000	03:37
Subject Sample	# 2	0.221	03:38
Breath Volume		2.339	03:38
Air Blank		0.000	03:39
Control Test		0.082	03:39
Air Blank		0.000	03:40
Diagnostics Check		OK	03:40

Data Download Date/Time: 01/24/2019 16:32

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HARWOOD,ARMANDO
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: PAREDES

DATE/TIME OF TEST: 01/23/2019 03:30:54

Diagnostics Check		OK	03:24
Air Blank		0.000	03:25
Control Test		0.082	03:25
Air Blank		0.000	03:25
Subject Sample	# 1	0.235	03:26
Breath Volume		3.859	03:26
Air Blank		0.000	03:26
Air Blank		0.000	03:28
Subject Sample	# 2	0.224	03:29
Breath Volume		4.062	03:29
Air Blank		0.000	03:30
Control Test		0.081	03:30
Air Blank		0.000	03:30
Diagnostics Check		OK	03:30

Data Download Date/Time: 01/24/2019 16:32

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ANDREWS,GARY
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 01/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 01/24/2019 01:32:45

Diagnostics Check		OK	01:26
Air Blank		0.000	01:27
Control Test		0.082	01:27
Air Blank		0.000	01:28
Subject Sample	# 1	0.095	01:28
Breath Volume		3.765	01:28
Air Blank		0.000	01:28
Air Blank		0.000	01:30
Subject Sample	# 2	0.098	01:31
Breath Volume		3.839	01:31
Air Blank		0.000	01:31
Control Test		0.081	01:32
Air Blank		0.000	01:32
Diagnostics Check		OK	01:32

Data Download Date/Time: 01/24/2019 16:32

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WARNER,MICHAEL
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 01/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: DICK

DATE/TIME OF TEST: 01/24/2019 03:20:35

Diagnostics Check		OK	03:16	
Air Blank		0.000	03:17	
Control Test		0.082	03:17	
Air Blank		0.000	03:18	
Subject Sample	# 1		03:19	Subject Test Refused
Breath Volume		0.000	03:19	
Air Blank		0.000	03:19	
Control Test		0.082	03:20	
Air Blank		0.000	03:20	
Diagnostics Check		OK	03:20	

Data Download Date/Time: 01/24/2019 16:32

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AZEVEDO,FABIO
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 01/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 01/25/2019 23:48:46

Diagnostics Check		OK	23:42
Air Blank		0.000	23:43
Control Test		0.082	23:43
Air Blank		0.000	23:43
Subject Sample	# 1	0.214	23:44
Breath Volume		1.449	23:44
Air Blank		0.000	23:45
Air Blank		0.000	23:46
Subject Sample	# 2	0.230	23:47
Breath Volume		1.222	23:47
Air Blank		0.000	23:47
Control Test		0.081	23:48
Air Blank		0.000	23:48
Diagnostics Check		OK	23:48

Data Download Date/Time: 01/31/2019 13:15

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HARRISON,ROBBIE
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: HO HONG

DATE/TIME OF TEST: 01/26/2019 01:57:05

Diagnostics Check		OK	01:45	
Air Blank		0.000	01:46	
Control Test		0.081	01:46	
Air Blank		0.000	01:47	
Subject Sample	# 1		01:50	No Sample Provided
Breath Volume		0.000	01:50	
Air Blank		0.000	01:50	
Air Blank		0.000	01:52	
Subject Sample	# 2		01:55	No Sample Provided
Breath Volume		0.000	01:55	
Air Blank		0.000	01:56	
Control Test		0.082	01:56	
Air Blank		0.000	01:57	
Diagnostics Check		OK	01:57	

Data Download Date/Time: 01/31/2019 13:15

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SOSA,EMMANUEL
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 01/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: VENOSE

DATE/TIME OF TEST: 01/26/2019 02:12:50

Diagnostics Check		OK	02:08	
Air Blank		0.000	02:08	
Control Test		0.082	02:08	
Air Blank		0.000	02:09	
Subject Sample	# 1		02:11	Subject Test Refused
Breath Volume		0.000	02:11	
Air Blank		0.000	02:11	
Control Test		0.081	02:12	
Air Blank		0.000	02:12	
Diagnostics Check		OK	02:12	

Data Download Date/Time: 01/31/2019 13:15

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CAMPBELL,CRAIG
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 01/26/2019 03:02:48

Diagnostics Check		OK	02:55
Air Blank		0.000	02:56
Control Test		0.082	02:56
Air Blank		0.000	02:57
Subject Sample	# 1	0.179	02:58
Breath Volume		1.277	02:58
Air Blank		0.000	02:58
Air Blank		0.000	03:00
Subject Sample	# 2	0.175	03:01
Breath Volume		1.539	03:01
Air Blank		0.000	03:01
Control Test		0.081	03:02
Air Blank		0.000	03:02
Diagnostics Check		OK	03:02

Data Download Date/Time: 01/31/2019 13:15

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MOONDA,SHUAIB
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 01/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 01/26/2019 04:29:02

Diagnostics Check		OK	04:22
Air Blank		0.000	04:23
Control Test		0.081	04:23
Air Blank		0.000	04:23
Subject Sample	# 1	0.166	04:24
Breath Volume		1.816	04:24
Air Blank		0.000	04:25
Air Blank		0.000	04:27
Subject Sample	# 2	0.163	04:27
Breath Volume		1.816	04:27
Air Blank		0.000	04:28
Control Test		0.081	04:28
Air Blank		0.000	04:28
Diagnostics Check		OK	04:29

Data Download Date/Time: 01/31/2019 13:15

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TOLIVER,JOVAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 01/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 01/26/2019 23:59:33

Diagnostics Check		OK	23:56	
Air Blank		0.000	23:56	
Control Test		0.081	23:56	
Air Blank		0.000	23:57	
Subject Sample	# 1		23:58	Subject Test Refused
Breath Volume		0.000	23:58	
Air Blank		0.000	23:58	
Control Test		0.082	23:58	
Air Blank		0.000	23:59	
Diagnostics Check		OK	23:59	

Data Download Date/Time: 01/31/2019 13:15

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HARRIS,BRIAN
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 01/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: CLAVER

DATE/TIME OF TEST: 01/27/2019 01:57:04

Diagnostics Check		OK	01:43	
Air Blank		0.000	01:43	
Control Test		0.081	01:43	
Air Blank		0.000	01:44	
Subject Sample	# 1	0.180	01:47	Volume Not Met
Breath Volume		0.292	01:47	
Air Blank		0.000	01:48	
Air Blank		0.000	01:49	
Subject Sample	# 2	0.196	01:52	
Breath Volume		1.175	01:52	
Air Blank		0.000	01:53	
Air Blank		0.000	01:55	
Subject Sample	# 3	0.197	01:55	
Breath Volume		1.761	01:55	
Air Blank		0.000	01:56	
Control Test		0.080	01:56	
Air Blank		0.000	01:56	
Diagnostics Check		OK	01:57	

Data Download Date/Time: 01/31/2019 13:15

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MONTGOMERY,JAMES
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: CARABALLO

DATE/TIME OF TEST: 01/27/2019 02:10:42

Diagnostics Check		OK	02:04
Air Blank		0.000	02:04
Control Test		0.081	02:05
Air Blank		0.000	02:05
Subject Sample	# 1	0.155	02:06
Breath Volume		2.585	02:06
Air Blank		0.000	02:06
Air Blank		0.000	02:08
Subject Sample	# 2	0.151	02:09
Breath Volume		1.519	02:09
Air Blank		0.000	02:09
Control Test		0.080	02:10
Air Blank		0.000	02:10
Diagnostics Check		OK	02:10

Data Download Date/Time: 01/31/2019 13:15

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CILENTI,A
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 01/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 01/27/2019 03:14:20

Diagnostics Check		OK	03:06
Air Blank		0.000	03:07
Control Test		0.081	03:07
Air Blank		0.000	03:08
Subject Sample	# 1	0.116	03:09
Breath Volume		1.671	03:09
Air Blank		0.000	03:09
Air Blank		0.000	03:11
Subject Sample	# 2	0.119	03:12
Breath Volume		2.355	03:12
Air Blank		0.000	03:13
Control Test		0.081	03:13
Air Blank		0.000	03:14
Diagnostics Check		OK	03:14

Data Download Date/Time: 01/31/2019 13:15

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANER,OSVALDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RITTER,ALEC
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: RESENDIZ-GUZMAN

DATE/TIME OF TEST: 01/28/2019 04:44:12

Diagnostics Check		OK	04:36
Air Blank		0.000	04:37
Control Test		0.081	04:37
Air Blank		0.000	04:38
Subject Sample	# 1	0.161	04:39
Breath Volume		1.761	04:39
Air Blank		0.000	04:39
Air Blank		0.000	04:41
Subject Sample	# 2	0.160	04:42
Breath Volume		2.328	04:42
Air Blank		0.000	04:43
Control Test		0.081	04:43
Air Blank		0.000	04:44
Diagnostics Check		OK	04:44

Data Download Date/Time: 01/31/2019 13:15

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HALLOCK,STEPHEN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: JAGGERS

DATE/TIME OF TEST: 01/30/2019 04:26:25

Diagnostics Check		OK	04:12	
Air Blank		0.000	04:13	
Control Test		0.081	04:13	
Air Blank		0.000	04:13	
Subject Sample	# 1	0.208	04:15	
Breath Volume		1.625	04:15	
Air Blank		0.000	04:16	
Air Blank		0.000	04:18	
Subject Sample	# 2	0.220	04:21	Volume Not Met
Breath Volume		0.824	04:21	
Air Blank		0.000	04:22	
Air Blank		0.000	04:23	
Subject Sample	# 3	0.197	04:24	
Breath Volume		2.125	04:24	
Air Blank		0.000	04:25	
Control Test		0.078	04:25	
Air Blank		0.000	04:26	
Diagnostics Check		OK	04:26	

Data Download Date/Time: 01/31/2019 13:15

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: COLE,KYLE
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: NASH

DATE/TIME OF TEST: 01/31/2019 02:51:59

Diagnostics Check		OK	02:49	
Air Blank		0.000	02:49	
Control Test		0.082	02:49	
Air Blank		0.000	02:50	
Subject Sample	# 1		02:50	Subject Test Refused
Breath Volume		0.000	02:50	
Air Blank		0.000	02:51	
Control Test		0.082	02:51	
Air Blank		0.000	02:51	
Diagnostics Check		OK	02:51	

Data Download Date/Time: 01/31/2019 13:15

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HARWOOD,ARMANDO
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: RIOS ESTEVES

DATE/TIME OF TEST: 01/31/2019 19:25:17

Diagnostics Check		OK	19:19
Air Blank		0.000	19:19
Control Test		0.081	19:20
Air Blank		0.000	19:20
Subject Sample	# 1	0.000	19:21
Breath Volume		2.152	19:21
Air Blank		0.000	19:21
Air Blank		0.000	19:23
Subject Sample	# 2	0.000	19:23
Breath Volume		3.179	19:23
Air Blank		0.000	19:24
Control Test		0.082	19:24
Air Blank		0.000	19:25
Diagnostics Check		OK	19:25

Data Download Date/Time: 02/07/2019 13:47

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MILAN,VICTOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:43

SUBJECT'S LAST NAME: VILLATORO

DATE/TIME OF TEST: 01/31/2019 21:11:31

Diagnostics Check		OK	21:05
Air Blank		0.000	21:05
Control Test		0.082	21:06
Air Blank		0.000	21:06
Subject Sample	# 1	0.146	21:07
Breath Volume		1.214	21:07
Air Blank		0.000	21:07
Air Blank		0.000	21:09
Subject Sample	# 2	0.146	21:10
Breath Volume		2.085	21:10
Air Blank		0.000	21:10
Control Test		0.081	21:10
Air Blank		0.000	21:11
Diagnostics Check		OK	21:11

Data Download Date/Time: 02/07/2019 13:47

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AZEVEDO,FABIO
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: DIXON

DATE/TIME OF TEST: 01/03/2019 04:10:47

Diagnostics Check		OK	04:07	
Air Blank		0.000	04:08	
Control Test		0.081	04:08	
Air Blank		0.000	04:09	
Subject Sample	# 1		04:09	Subject Test Refused
Breath Volume		0.000	04:09	
Air Blank		0.000	04:09	
Control Test		0.079	04:10	
Air Blank		0.000	04:10	
Diagnostics Check		OK	04:10	

Data Download Date/Time: 01/03/2019 13:50

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HEBEL,GRIFFIN
ARRESTING OFFICER AGENCY: Windermere PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 01/05/2019 04:02:42

Diagnostics Check		OK	03:56
Air Blank		0.000	03:56
Control Test		0.081	03:56
Air Blank		0.000	03:57
Subject Sample	# 1	0.199	03:57
Breath Volume		1.578	03:57
Air Blank		0.000	03:58
Air Blank		0.000	04:00
Subject Sample	# 2	0.189	04:00
Breath Volume		1.238	04:00
Air Blank		0.000	04:01
Control Test		0.081	04:02
Air Blank		0.000	04:02
Diagnostics Check		OK	04:02

Data Download Date/Time: 01/16/2019 14:53

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAWKINS,DEBORAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: AMADOR

DATE/TIME OF TEST: 01/06/2019 02:46:08

Diagnostics Check		OK	02:39
Air Blank		0.000	02:40
Control Test		0.081	02:40
Air Blank		0.000	02:41
Subject Sample	# 1	0.149	02:41
Breath Volume		1.753	02:41
Air Blank		0.000	02:42
Air Blank		0.000	02:44
Subject Sample	# 2	0.151	02:44
Breath Volume		2.050	02:44
Air Blank		0.000	02:45
Control Test		0.081	02:45
Air Blank		0.000	02:46
Diagnostics Check		OK	02:46

Data Download Date/Time: 01/16/2019 14:53

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KELEMEN,JANOS
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 01/09/2019 03:37:27

Diagnostics Check		OK	03:30
Air Blank		0.000	03:30
Control Test		0.080	03:30
Air Blank		0.000	03:31
Subject Sample	# 1	0.183	03:32
Breath Volume		1.796	03:32
Air Blank		0.000	03:33
Air Blank		0.000	03:34
Subject Sample	# 2	0.185	03:35
Breath Volume		1.324	03:35
Air Blank		0.000	03:36
Control Test		0.080	03:36
Air Blank		0.000	03:37
Diagnostics Check		OK	03:37

Data Download Date/Time: 01/16/2019 14:53

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VAUGHN,JACOB
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 01/12/2019 03:53:01

Diagnostics Check		OK	03:46
Air Blank		0.000	03:47
Control Test		0.081	03:47
Air Blank		0.000	03:47
Subject Sample	# 1	0.095	03:48
Breath Volume		2.320	03:48
Air Blank		0.000	03:49
Air Blank		0.000	03:51
Subject Sample	# 2	0.095	03:51
Breath Volume		2.054	03:51
Air Blank		0.000	03:52
Control Test		0.082	03:52
Air Blank		0.000	03:52
Diagnostics Check		OK	03:53

Data Download Date/Time: 01/16/2019 14:53

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HOLIDAY,LUCAS
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 01/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: VALENTIN

DATE/TIME OF TEST: 01/17/2019 01:38:21

Diagnostics Check		OK	01:31
Air Blank		0.000	01:32
Control Test		0.082	01:32
Air Blank		0.000	01:33
Subject Sample	# 1	0.000	01:33
Breath Volume		1.742	01:33
Air Blank		0.000	01:34
Air Blank		0.000	01:36
Subject Sample	# 2	0.000	01:36
Breath Volume		1.992	01:36
Air Blank		0.000	01:37
Control Test		0.082	01:37
Air Blank		0.000	01:38
Diagnostics Check		OK	01:38

Data Download Date/Time: 01/24/2019 16:25

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAWKINS,DEBBIE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: NEWHAM

DATE/TIME OF TEST: 01/18/2019 03:50:19

Diagnostics Check		OK	03:43
Air Blank		0.000	03:44
Control Test		0.081	03:44
Air Blank		0.000	03:45
Subject Sample	# 1	0.186	03:45
Breath Volume		1.664	03:45
Air Blank		0.000	03:46
Air Blank		0.000	03:48
Subject Sample	# 2	0.205	03:48
Breath Volume		4.425	03:48
Air Blank		0.000	03:49
Control Test		0.081	03:49
Air Blank		0.000	03:50
Diagnostics Check		OK	03:50

Data Download Date/Time: 01/24/2019 16:25

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DAVILA,SAMUEL
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 01/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: MON

DATE/TIME OF TEST: 01/19/2019 04:21:08

Diagnostics Check		OK	04:14
Air Blank		0.000	04:14
Control Test		0.082	04:15
Air Blank		0.000	04:15
Subject Sample	# 1	0.159	04:16
Breath Volume		3.328	04:16
Air Blank		0.000	04:16
Air Blank		0.000	04:18
Subject Sample	# 2	0.158	04:19
Breath Volume		2.222	04:19
Air Blank		0.000	04:20
Control Test		0.081	04:20
Air Blank		0.000	04:21
Diagnostics Check		OK	04:21

Data Download Date/Time: 01/24/2019 16:25

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LEE II,MIKEL
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 01/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: BOURNES

DATE/TIME OF TEST: 01/23/2019 03:51:17

Diagnostics Check		OK	03:44
Air Blank		0.000	03:44
Control Test		0.081	03:44
Air Blank		0.000	03:45
Subject Sample	# 1	0.184	03:46
Breath Volume		2.835	03:46
Air Blank		0.000	03:46
Air Blank		0.000	03:48
Subject Sample	# 2	0.174	03:49
Breath Volume		2.503	03:49
Air Blank		0.000	03:50
Control Test		0.082	03:50
Air Blank		0.000	03:51
Diagnostics Check		OK	03:51

Data Download Date/Time: 01/24/2019 16:25

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MARTINEZ,ERNESTO
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: SIERRA

DATE/TIME OF TEST: 01/24/2019 01:51:01

Diagnostics Check		OK	01:44
Air Blank		0.000	01:45
Control Test		0.081	01:45
Air Blank		0.000	01:45
Subject Sample	# 1	0.000	01:46
Breath Volume		2.871	01:46
Air Blank		0.000	01:47
Air Blank		0.000	01:49
Subject Sample	# 2	0.000	01:49
Breath Volume		2.492	01:49
Air Blank		0.000	01:50
Control Test		0.081	01:50
Air Blank		0.000	01:50
Diagnostics Check		OK	01:51

Data Download Date/Time: 01/24/2019 16:25

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KELEMEN,JANOS
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: LONG

DATE/TIME OF TEST: 01/30/2019 03:46:50

Diagnostics Check		OK	03:40
Air Blank		0.000	03:40
Control Test		0.081	03:40
Air Blank		0.000	03:41
Subject Sample	# 1	0.149	03:42
Breath Volume		1.945	03:42
Air Blank		0.000	03:42
Air Blank		0.000	03:44
Subject Sample	# 2	0.146	03:45
Breath Volume		2.730	03:45
Air Blank		0.000	03:45
Control Test		0.082	03:46
Air Blank		0.000	03:46
Diagnostics Check		OK	03:46

Data Download Date/Time: 01/31/2019 13:12

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MARTINEZ,ERNESTO
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: JOSEPH

DATE/TIME OF TEST: 01/31/2019 03:13:31

Diagnostics Check		OK	03:06
Air Blank		0.000	03:06
Control Test		0.082	03:07
Air Blank		0.000	03:07
Subject Sample	# 1	0.115	03:09
Breath Volume		2.664	03:09
Air Blank		0.000	03:09
Air Blank		0.000	03:11
Subject Sample	# 2	0.108	03:11
Breath Volume		1.445	03:11
Air Blank		0.000	03:12
Control Test		0.081	03:12
Air Blank		0.000	03:13
Diagnostics Check		OK	03:13

Data Download Date/Time: 01/31/2019 13:12

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MILLS,RESHARD
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: SNIDER

DATE/TIME OF TEST: 01/31/2019 23:20:50

Diagnostics Check		OK	23:14
Air Blank		0.000	23:14
Control Test		0.081	23:15
Air Blank		0.000	23:15
Subject Sample	# 1	0.231	23:16
Breath Volume		2.593	23:16
Air Blank		0.000	23:16
Air Blank		0.000	23:18
Subject Sample	# 2	0.217	23:19
Breath Volume		1.515	23:19
Air Blank		0.000	23:19
Control Test		0.080	23:20
Air Blank		0.000	23:20
Diagnostics Check		OK	23:20

Data Download Date/Time: 02/13/2019 11:43

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARLOW,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WHITE,MYRON
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:23

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 01/04/2019 05:49:51

Diagnostics Check		OK	05:43
Air Blank		0.000	05:43
Control Test		0.081	05:44
Air Blank		0.000	05:44
Subject Sample	# 1	0.156	05:45
Breath Volume		2.996	05:45
Air Blank		0.000	05:45
Air Blank		0.000	05:47
Subject Sample	# 2	0.156	05:48
Breath Volume		2.828	05:48
Air Blank		0.000	05:48
Control Test		0.081	05:49
Air Blank		0.000	05:49
Diagnostics Check		OK	05:49

Data Download Date/Time: 01/11/2019 16:21

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTSTATT,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: CASSANI,D
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 01/06/2019 02:20:38

Diagnostics Check		OK	02:09	
Air Blank		0.000	02:09	
Control Test		0.081	02:09	
Air Blank		0.000	02:10	
Subject Sample	# 1	0.166	02:13	Volume Not Met
Breath Volume		0.707	02:13	
Air Blank		0.000	02:14	
Air Blank		0.000	02:15	
Subject Sample	# 2	0.190	02:19	Volume Not Met
Breath Volume		0.234	02:19	
Air Blank		0.000	02:19	
Control Test		0.080	02:20	
Air Blank		0.000	02:20	
Diagnostics Check		OK	02:20	

Data Download Date/Time: 01/11/2019 16:21

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARLOW,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: BAHNSEN,ANDREW
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: RIVERS

DATE/TIME OF TEST: 01/09/2019 01:11:36

Diagnostics Check		OK	01:00	
Air Blank		0.000	01:00	
Control Test		0.081	01:01	
Air Blank		0.000	01:01	
Subject Sample	# 1		01:04	No Sample Provided
Breath Volume		0.000	01:04	
Air Blank		0.000	01:05	
Air Blank		0.000	01:07	
Subject Sample	# 2		01:10	No Sample Provided
Breath Volume		0.000	01:10	
Air Blank		0.000	01:10	
Control Test		0.081	01:11	
Air Blank		0.000	01:11	
Diagnostics Check		OK	01:11	

Data Download Date/Time: 01/11/2019 16:21

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENTZ,JARROD,C
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: STIVERS,H
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: TOMAZINIS

DATE/TIME OF TEST: 01/13/2019 01:38:19

Diagnostics Check		OK	01:35	
Air Blank		0.000	01:35	
Control Test		0.081	01:35	
Air Blank		0.000	01:36	
Subject Sample	# 1		01:36	Subject Test Refused
Breath Volume		0.000	01:36	
Air Blank		0.000	01:37	
Control Test		0.081	01:37	
Air Blank		0.000	01:38	
Diagnostics Check		OK	01:38	

Data Download Date/Time: 02/22/2019 11:39

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARLOW,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: PHILLIPS,CHARLIE
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 01/13/2019 03:59:19

Diagnostics Check		OK	03:50	
Air Blank		0.000	03:50	
Control Test		0.081	03:51	
Air Blank		0.000	03:51	
Subject Sample	# 1		03:54	No Sample Provided
Breath Volume		0.000	03:54	
Air Blank		0.000	03:55	
Air Blank		0.000	03:57	
Subject Sample	# 2		03:58	Subject Test Refused
Breath Volume		0.000	03:58	
Air Blank			03:58	RFI Detect
Air Blank		0.000	03:59	

Data Download Date/Time: 02/22/2019 11:39

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARLOW,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: PHILLIPS,CHARLIE
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 01/13/2019 04:12:15

Diagnostics Check		OK	04:03	
Air Blank		0.000	04:04	
Control Test		0.081	04:04	
Air Blank		0.000	04:05	
Subject Sample	# 1		04:08	No Sample Provided
Breath Volume		0.000	04:08	
Air Blank		0.000	04:08	
Air Blank		0.000	04:10	
Subject Sample	# 2		04:10	Subject Test Refused
Breath Volume		0.000	04:10	
Air Blank		0.000	04:11	
Control Test		0.081	04:11	
Air Blank		0.000	04:12	
Diagnostics Check		OK	04:12	

Data Download Date/Time: 02/22/2019 11:39

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARCHER,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GONZALEZ,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 12:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:38

SUBJECT'S LAST NAME: MOCK

DATE/TIME OF TEST: 01/19/2019 15:13:29

Diagnostics Check		OK	15:06
Air Blank		0.000	15:06
Control Test		0.081	15:07
Air Blank		0.000	15:07
Subject Sample	# 1	0.260	15:09
Breath Volume		1.460	15:09
Air Blank		0.000	15:09
Air Blank		0.000	15:11
Subject Sample	# 2	0.269	15:12
Breath Volume		1.921	15:12
Air Blank		0.000	15:12
Control Test		0.080	15:12
Air Blank		0.000	15:13
Diagnostics Check		OK	15:13

Data Download Date/Time: 02/22/2019 11:39

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTIAGO,DEPUTY,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: PRITCHARD,DEPUTY
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:25

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 01/21/2019 17:55:53

Diagnostics Check		OK	17:49
Air Blank		0.000	17:49
Control Test		0.082	17:50
Air Blank		0.000	17:50
Subject Sample	# 1	0.057	17:51
Breath Volume		1.601	17:51
Air Blank		0.000	17:51
Air Blank		0.000	17:53
Subject Sample	# 2	0.056	17:54
Breath Volume		1.468	17:54
Air Blank		0.000	17:54
Control Test		0.081	17:55
Air Blank		0.000	17:55
Diagnostics Check		OK	17:55

Data Download Date/Time: 02/22/2019 11:39

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001646

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURPHY,PATRICK,J
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: TESTER,TEST
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 01/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 08:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:30

SUBJECT'S LAST NAME: TESTER

DATE/TIME OF TEST: 01/03/2019 09:52:01

Diagnostics Check	OK	09:50	
Air Blank	0.000	09:51	
Control Test	0.000	09:51	Control Outside Tolerance
Air Blank	0.000	09:52	

Data Download Date/Time: 01/04/2019 11:47

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FILLAK,MICHAEL,O
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: LEONARD,STEPHEN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: GIBSON

DATE/TIME OF TEST: 01/03/2019 01:46:58

Diagnostics Check		OK	01:33	
Air Blank		0.000	01:34	
Control Test		0.080	01:34	
Air Blank		0.000	01:34	
Subject Sample	# 1	0.189	01:36	
Breath Volume		2.632	01:36	
Air Blank		0.000	01:37	
Air Blank		0.000	01:39	
Subject Sample	# 2	0.170	01:42	Slope Not Level
Breath Volume		2.691	01:42	
Air Blank		0.000	01:42	
Air Blank		0.000	01:44	
Subject Sample	# 3	0.187	01:45	
Breath Volume		2.839	01:45	
Air Blank		0.000	01:46	
Control Test		0.079	01:46	
Air Blank		0.000	01:46	
Diagnostics Check		OK	01:46	

Data Download Date/Time: 01/09/2019 11:15

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINI,RYAN,H
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MARTINI,RYAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 0521808A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:29

SUBJECT'S LAST NAME: REAGAN

DATE/TIME OF TEST: 01/03/2019 19:58:31

Diagnostics Check		OK	19:52
Air Blank		0.000	19:53
Control Test		0.080	19:53
Air Blank		0.000	19:53
Subject Sample	# 1	0.000	19:54
Breath Volume		2.851	19:54
Air Blank		0.000	19:54
Air Blank		0.000	19:56
Subject Sample	# 2	0.000	19:57
Breath Volume		2.175	19:57
Air Blank		0.000	19:57
Control Test		0.079	19:57
Air Blank		0.000	19:58
Diagnostics Check		OK	19:58

Data Download Date/Time: 01/09/2019 11:15

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINI,RYAN,H
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MARTINI,RYAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 01/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: PETERS

DATE/TIME OF TEST: 01/13/2019 20:45:34

Diagnostics Check		OK	20:35
Air Blank		0.000	20:36
Control Test		0.080	20:36
Air Blank		0.000	20:37
Subject Sample	# 1	0.356	20:38
Breath Volume		2.292	20:38
Air Blank		0.000	20:38
Air Blank		0.000	20:40
Subject Sample	# 2	0.329	20:41
Breath Volume		2.328	20:41
Air Blank		0.000	20:41
Air Blank		0.000	20:43
Subject Sample	# 3	0.345	20:44
Breath Volume		2.390	20:44
Air Blank		0.000	20:44
Control Test		0.078	20:44
Air Blank		0.000	20:45
Diagnostics Check		OK	20:45

Data Download Date/Time: 02/15/2019 09:47

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,PERRY,M
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: WILSON,PERRY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 01/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: MORRISON

DATE/TIME OF TEST: 01/14/2019 02:03:08

Diagnostics Check		OK	01:55
Air Blank		0.000	01:56
Control Test		0.079	01:56
Air Blank		0.000	01:57
Subject Sample	# 1	0.169	01:58
Breath Volume		2.261	01:58
Air Blank		0.000	01:59
Air Blank		0.000	02:01
Subject Sample	# 2	0.167	02:01
Breath Volume		1.960	02:01
Air Blank		0.000	02:02
Control Test		0.078	02:02
Air Blank		0.000	02:03
Diagnostics Check		OK	02:03

Data Download Date/Time: 02/15/2019 09:47

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,ALFONSO,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: RODRIGUEZ,ALFONSO
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 01/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: SNELL

DATE/TIME OF TEST: 01/15/2019 04:10:11

Diagnostics Check		OK	04:03
Air Blank		0.000	04:03
Control Test		0.080	04:04
Air Blank		0.000	04:04
Subject Sample	# 1	0.047	04:05
Breath Volume		2.371	04:05
Air Blank		0.000	04:05
Air Blank		0.000	04:07
Subject Sample	# 2	0.050	04:08
Breath Volume		1.781	04:08
Air Blank		0.000	04:09
Control Test		0.080	04:09
Air Blank		0.000	04:10
Diagnostics Check		OK	04:10

Data Download Date/Time: 02/15/2019 09:48

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,ALFONSO,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: THOMAS,DANIEL
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 01/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: DANIELSON-WALNY

DATE/TIME OF TEST: 01/16/2019 00:52:44

Diagnostics Check		OK	00:46
Air Blank		0.000	00:47
Control Test		0.080	00:47
Air Blank		0.000	00:47
Subject Sample	# 1	0.000	00:48
Breath Volume		3.109	00:48
Air Blank		0.000	00:48
Air Blank		0.000	00:50
Subject Sample	# 2	0.000	00:51
Breath Volume		3.089	00:51
Air Blank		0.000	00:51
Control Test		0.079	00:52
Air Blank		0.000	00:52
Diagnostics Check		OK	00:52

Data Download Date/Time: 02/15/2019 09:48

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUFFY,J,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: DUFFY,J
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 01/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:51

SUBJECT'S LAST NAME: BORECKI

DATE/TIME OF TEST: 01/19/2019 00:17:22

Diagnostics Check		OK	00:11	
Air Blank		0.000	00:12	
Control Test		0.080	00:12	
Air Blank		0.000	00:13	
Subject Sample	# 1		00:16	Subject Test Refused
Breath Volume		0.000	00:16	
Air Blank		0.000	00:16	
Control Test		0.080	00:16	
Air Blank		0.000	00:17	
Diagnostics Check		OK	00:17	

Data Download Date/Time: 02/15/2019 09:48

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUFFY,J,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: HARRISON,W
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 01/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: GONZALEZ SABORIT

DATE/TIME OF TEST: 01/19/2019 04:33:33

Diagnostics Check		OK	04:22	
Air Blank		0.000	04:22	
Control Test		0.080	04:22	
Air Blank		0.000	04:23	
Subject Sample	# 1		04:26	No Sample Provided
Breath Volume		0.000	04:26	
Air Blank		0.000	04:27	
Air Blank		0.000	04:29	
Subject Sample	# 2		04:32	No Sample Provided
Breath Volume		0.000	04:32	
Air Blank		0.000	04:32	
Control Test		0.079	04:32	
Air Blank		0.000	04:33	
Diagnostics Check		OK	04:33	

Data Download Date/Time: 02/15/2019 09:48

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUSKEVICH,CURT,E
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: SUSKEVICH,C
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 01/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: ALFONSO

DATE/TIME OF TEST: 01/20/2019 02:29:02

Diagnostics Check		OK	02:21
Air Blank		0.000	02:21
Control Test		0.079	02:21
Air Blank		0.000	02:22
Subject Sample	# 1	0.177	02:24
Breath Volume		1.457	02:24
Air Blank		0.000	02:24
Air Blank		0.000	02:26
Subject Sample	# 2	0.166	02:27
Breath Volume		1.500	02:27
Air Blank		0.000	02:28
Control Test		0.079	02:28
Air Blank		0.000	02:28
Diagnostics Check		OK	02:29

Data Download Date/Time: 02/15/2019 09:48

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINI,RYAN,H
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MADURO,VICTOR
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 01/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:54

SUBJECT'S LAST NAME: HOFFMAN

DATE/TIME OF TEST: 01/22/2019 19:20:18

Diagnostics Check		OK	19:14
Air Blank		0.000	19:14
Control Test		0.080	19:15
Air Blank		0.000	19:15
Subject Sample	# 1	0.145	19:15
Breath Volume		3.707	19:15
Air Blank		0.000	19:16
Air Blank		0.000	19:18
Subject Sample	# 2	0.152	19:18
Breath Volume		3.089	19:18
Air Blank		0.000	19:19
Control Test		0.079	19:19
Air Blank		0.000	19:20
Diagnostics Check		OK	19:20

Data Download Date/Time: 02/15/2019 09:48

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FESSEL,DANIEL,C
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MONTGOMERY,JOHN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 01/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 15:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:12

SUBJECT'S LAST NAME: NAVARRO

DATE/TIME OF TEST: 01/28/2019 16:46:04

Diagnostics Check		OK	16:39
Air Blank		0.000	16:40
Control Test		0.079	16:40
Air Blank		0.000	16:41
Subject Sample	# 1	0.192	16:41
Breath Volume		1.429	16:41
Air Blank		0.000	16:42
Air Blank		0.000	16:43
Subject Sample	# 2	0.199	16:44
Breath Volume		2.800	16:44
Air Blank		0.000	16:45
Control Test		0.079	16:45
Air Blank		0.000	16:45
Diagnostics Check		OK	16:46

Data Download Date/Time: 02/15/2019 09:48

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUFFY,J,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: THOMAS,D
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 01/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:43

SUBJECT'S LAST NAME: SCHWANDERMAN

DATE/TIME OF TEST: 01/29/2019 20:20:21

Diagnostics Check		OK	20:09	
Air Blank		0.000	20:09	
Control Test		0.079	20:09	
Air Blank		0.000	20:10	
Subject Sample	# 1		20:13	No Sample Provided
Breath Volume		0.000	20:13	
Air Blank		0.000	20:13	
Air Blank		0.000	20:15	
Subject Sample	# 2		20:19	No Sample Provided
Breath Volume		0.000	20:19	
Air Blank		0.000	20:19	
Control Test		0.080	20:19	
Air Blank		0.000	20:20	
Diagnostics Check		OK	20:20	

Data Download Date/Time: 02/15/2019 09:48

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,JOSE,D
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: LOPEZ,JOSE
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 01/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: TORRES IXCOY

DATE/TIME OF TEST: 01/13/2019 03:12:14

Diagnostics Check		OK	03:05
Air Blank		0.000	03:06
Control Test		0.078	03:06
Air Blank		0.000	03:06
Subject Sample	# 1	0.231	03:07
Breath Volume		2.585	03:07
Air Blank		0.000	03:08
Air Blank		0.000	03:09
Subject Sample	# 2	0.224	03:10
Breath Volume		1.945	03:10
Air Blank		0.000	03:11
Control Test		0.078	03:11
Air Blank		0.000	03:12
Diagnostics Check		OK	03:12

Data Download Date/Time: 02/04/2019 10:24

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLELLAND,WILLIAM,R
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MORGAN,J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 01/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: ARROYO

DATE/TIME OF TEST: 01/19/2019 03:37:37

Diagnostics Check		OK	03:31
Air Blank		0.000	03:31
Control Test		0.078	03:32
Air Blank		0.000	03:32
Subject Sample	# 1	0.179	03:33
Breath Volume		2.886	03:33
Air Blank		0.000	03:33
Air Blank		0.000	03:35
Subject Sample	# 2	0.176	03:36
Breath Volume		3.113	03:36
Air Blank		0.000	03:36
Control Test		0.078	03:36
Air Blank		0.000	03:37
Diagnostics Check		OK	03:37

Data Download Date/Time: 02/04/2019 10:24

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,JOSE,D
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: LOPEZ,JOSE
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 01/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: ESTRADA

DATE/TIME OF TEST: 01/27/2019 01:40:59

Diagnostics Check		OK	01:34
Air Blank		0.000	01:35
Control Test		0.079	01:35
Air Blank		0.000	01:35
Subject Sample	# 1	0.164	01:36
Breath Volume		2.027	01:36
Air Blank		0.000	01:37
Air Blank		0.000	01:38
Subject Sample	# 2	0.172	01:39
Breath Volume		1.441	01:39
Air Blank		0.000	01:40
Control Test		0.079	01:40
Air Blank		0.000	01:40
Diagnostics Check		OK	01:40

Data Download Date/Time: 02/04/2019 10:24

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CURTO,THEODORE,D
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MORERA,D
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 01/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: MENDOZADAVALOS

DATE/TIME OF TEST: 01/27/2019 03:53:06

Diagnostics Check		OK	03:46
Air Blank		0.000	03:46
Control Test		0.079	03:47
Air Blank		0.000	03:47
Subject Sample	# 1	0.247	03:48
Breath Volume		2.609	03:48
Air Blank		0.000	03:48
Air Blank		0.000	03:50
Subject Sample	# 2	0.257	03:51
Breath Volume		2.289	03:51
Air Blank		0.000	03:52
Control Test		0.077	03:52
Air Blank		0.000	03:53
Diagnostics Check		OK	03:53

Data Download Date/Time: 02/04/2019 10:24

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUM,ZACKARY,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WOOLVERTON,JUSTIN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: VAUGHAN

DATE/TIME OF TEST: 01/10/2019 03:36:35

Diagnostics Check		OK	03:28
Air Blank		0.000	03:29
Control Test		0.079	03:29
Air Blank		0.000	03:29
Subject Sample	# 1	0.298	03:31
Breath Volume		1.527	03:31
Air Blank		0.000	03:32
Air Blank		0.000	03:33
Subject Sample	# 2	0.309	03:35
Breath Volume		1.921	03:35
Air Blank		0.000	03:35
Control Test		0.079	03:35
Air Blank		0.000	03:36
Diagnostics Check		OK	03:36

Data Download Date/Time: 01/15/2019 08:36

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,S,T
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GIANNOTTA,D
ARRESTING OFFICER AGENCY: St. Augustine Beach
LAST AGENCY INSPECTION: 01/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 01/20/2019 03:28:17

Diagnostics Check		OK	03:11	
Air Blank		0.000	03:11	
Control Test		0.082	03:12	
Air Blank		0.000	03:12	
Subject Sample	# 1	0.160	03:15	
Breath Volume		1.277	03:15	
Air Blank		0.000	03:15	
Air Blank		0.000	03:17	
Subject Sample	# 2	0.142	03:20	Volume Not Met
Breath Volume		0.066	03:20	
Air Blank		0.000	03:21	
Air Blank		0.000	03:23	
Subject Sample	# 3	0.135	03:26	Volume Not Met
Breath Volume		0.191	03:26	
Air Blank		0.000	03:27	
Control Test		0.079	03:27	
Air Blank		0.000	03:28	
Diagnostics Check		OK	03:28	

Data Download Date/Time: 02/26/2019 14:35

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BALLOUN,SAMUEL,C
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: GREEN,INDIA
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 01/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: SHANNON

DATE/TIME OF TEST: 01/24/2019 03:03:31

Diagnostics Check		OK	02:57
Air Blank		0.000	02:57
Control Test		0.082	02:57
Air Blank		0.000	02:58
Subject Sample	# 1	0.327	02:59
Breath Volume		2.691	02:59
Air Blank		0.000	02:59
Air Blank		0.000	03:01
Subject Sample	# 2	0.310	03:02
Breath Volume		2.777	03:02
Air Blank		0.000	03:02
Control Test		0.081	03:02
Air Blank		0.000	03:03
Diagnostics Check		OK	03:03

Data Download Date/Time: 02/26/2019 14:35

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001656

INSTRUMENT REGISTERED TO: NORTH MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PINEDA,MIGUEL,A
BREATH TEST OPERATOR AGENCY: NORTH MIAMI PD
ARRESTING OFFICER: PINEDA,MIGUEL
ARRESTING OFFICER AGENCY: NORTH MIAMI PD
LAST AGENCY INSPECTION: 01/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: STANLEY

DATE/TIME OF TEST: 01/13/2019 01:14:16

Diagnostics Check		OK	01:00	
Air Blank		0.000	01:01	
Control Test		0.081	01:01	
Air Blank		0.000	01:01	
Subject Sample	# 1	0.150	01:05	Volume Not Met
Breath Volume		1.003	01:05	
Air Blank		0.000	01:05	
Air Blank		0.000	01:07	
Subject Sample	# 2	0.154	01:08	
Breath Volume		1.507	01:08	
Air Blank		0.000	01:08	
Air Blank		0.000	01:10	
Subject Sample	# 3	0.153	01:12	
Breath Volume		1.515	01:12	
Air Blank		0.000	01:13	
Control Test		0.080	01:13	
Air Blank		0.000	01:14	
Diagnostics Check		OK	01:14	

Data Download Date/Time: 02/07/2019 12:18

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001656

INSTRUMENT REGISTERED TO: NORTH MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PINEDA,MIGUEL,A
BREATH TEST OPERATOR AGENCY: NORTH MIAMI PD
ARRESTING OFFICER: MURPHY,ANTHONY
ARRESTING OFFICER AGENCY: NORTH MIAMI PD
LAST AGENCY INSPECTION: 01/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: SARRAS BUTRON

DATE/TIME OF TEST: 01/13/2019 04:23:06

Diagnostics Check		OK	04:19	
Air Blank		0.000	04:20	
Control Test		0.080	04:20	
Air Blank		0.000	04:21	
Subject Sample	# 1		04:21	Subject Test Refused
Breath Volume		0.000	04:21	
Air Blank		0.000	04:22	
Control Test		0.080	04:22	
Air Blank		0.000	04:23	
Diagnostics Check		OK	04:23	

Data Download Date/Time: 02/07/2019 12:18

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001657

INSTRUMENT REGISTERED TO: MICCOSUKEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVAS-CUADRA,NORMAN,E
BREATH TEST OPERATOR AGENCY: Miccosukee PD
ARRESTING OFFICER: LIPARI,M
ARRESTING OFFICER AGENCY: Miccosukee PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A2
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 05:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:53

SUBJECT'S LAST NAME: RIVAS-CUADRA

DATE/TIME OF TEST: 01/07/2019 06:21:31

Diagnostics Check		OK	06:14
Air Blank		0.000	06:15
Control Test		0.077	06:15
Air Blank		0.000	06:16
Subject Sample	# 1	0.192	06:16
Breath Volume		1.550	06:16
Air Blank		0.000	06:17
Air Blank		0.000	06:19
Subject Sample	# 2	0.188	06:19
Breath Volume		1.707	06:19
Air Blank		0.000	06:20
Control Test		0.077	06:20
Air Blank		0.000	06:21
Diagnostics Check		OK	06:21

Data Download Date/Time: 01/10/2019 12:49

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARTANOWICZ,MARTIN,
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: KARTANOWICZ,MARTIN
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 01/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 11:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: MANNING

DATE/TIME OF TEST: 01/11/2019 01:05:35

Diagnostics Check	OK	01:04	
Air Blank	0.000	01:04	
Control Test	0.079	01:05	RFI Detect
Air Blank	0.000	01:05	

Data Download Date/Time: 01/24/2019 22:17

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARTANOWICZ,MARTIN,
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: KARTANOWICZ,MARTIN
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 01/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 11:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: MANNING

DATE/TIME OF TEST: 01/11/2019 01:28:09

Diagnostics Check	OK	01:27	
Air Blank		01:27	Ambient Fail
Air Blank	0.000	01:28	

Data Download Date/Time: 01/24/2019 22:17

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARTANOWICZ,MARTIN,
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: KARTANOWICZ,M
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 01/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 11:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: MANNING

DATE/TIME OF TEST: 01/11/2019 01:59:58

Diagnostics Check		OK	01:54	
Air Blank		0.000	01:54	
Control Test		0.079	01:55	
Air Blank		0.000	01:55	
Subject Sample	# 1	0.097	01:56	
Breath Volume		1.457	01:56	
Air Blank		0.000	01:57	
Air Blank		0.000	01:58	
Subject Sample	# 2	0.095	01:59	
Breath Volume		1.902	01:59	
Air Blank			01:59	RFI Detect
Air Blank		0.000	01:59	

Data Download Date/Time: 01/24/2019 22:17

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAYLE,OWEN,D
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: ROMAN,LEOPOLDO
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 01/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 12:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:19

SUBJECT'S LAST NAME: JOHNS

DATE/TIME OF TEST: 01/13/2019 13:45:44

Diagnostics Check		OK	13:39
Air Blank		0.000	13:40
Control Test		0.080	13:40
Air Blank		0.000	13:41
Subject Sample	# 1	0.000	13:41
Breath Volume		1.789	13:41
Air Blank		0.000	13:41
Air Blank		0.000	13:43
Subject Sample	# 2	0.000	13:44
Breath Volume		2.113	13:44
Air Blank		0.000	13:44
Control Test		0.078	13:45
Air Blank		0.000	13:45
Diagnostics Check		OK	13:45

Data Download Date/Time: 01/24/2019 22:17

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAYERS,MICHELLE,H
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: SAYERS,MICHELLE
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 01/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: BRANSTETTER

DATE/TIME OF TEST: 01/15/2019 00:46:03

Diagnostics Check		OK	00:39
Air Blank		0.000	00:39
Control Test		0.080	00:39
Air Blank		0.000	00:40
Subject Sample	# 1	0.153	00:41
Breath Volume		1.218	00:41
Air Blank		0.000	00:42
Air Blank		0.000	00:43
Subject Sample	# 2	0.156	00:44
Breath Volume		1.777	00:44
Air Blank		0.000	00:45
Control Test		0.080	00:45
Air Blank		0.000	00:45
Diagnostics Check		OK	00:46

Data Download Date/Time: 01/24/2019 22:17

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TELLERIA,TROY,R
BREATH TEST OPERATOR AGENCY: St. Cloud PD
ARRESTING OFFICER: TELLERIA,TROY
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 01/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:37

SUBJECT'S LAST NAME: HOOVER

DATE/TIME OF TEST: 01/23/2019 21:19:35

Diagnostics Check		OK	21:11
Air Blank		0.000	21:12
Control Test		0.080	21:12
Air Blank		0.000	21:12
Subject Sample	# 1	0.206	21:14
Breath Volume		1.597	21:14
Air Blank		0.000	21:14
Air Blank		0.000	21:16
Subject Sample	# 2	0.210	21:17
Breath Volume		2.835	21:17
Air Blank		0.000	21:18
Control Test		0.078	21:18
Air Blank		0.000	21:19
Diagnostics Check		OK	21:19

Data Download Date/Time: 01/24/2019 22:17

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001720

INSTRUMENT REGISTERED TO: ALACHUA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKS,DONALD,M
BREATH TEST OPERATOR AGENCY: Alachua PD
ARRESTING OFFICER: HUGHES,DAVID
ARRESTING OFFICER AGENCY: Alachua PD
LAST AGENCY INSPECTION: 01/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 805701
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 05:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:50

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 01/17/2019 06:27:03

Diagnostics Check		OK	06:19
Air Blank		0.000	06:20
Control Test		0.080	06:20
Air Blank		0.000	06:21
Subject Sample	# 1	0.120	06:22
Breath Volume		2.113	06:22
Air Blank		0.000	06:23
Air Blank		0.000	06:24
Subject Sample	# 2	0.109	06:25
Breath Volume		1.164	06:25
Air Blank		0.000	06:26
Control Test		0.078	06:26
Air Blank		0.000	06:26
Diagnostics Check		OK	06:27

Data Download Date/Time: 02/24/2019 23:38

Printed: 10 April 2019

COBRA ©1997-2014 CMI, Inc.