Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGBY,CHARLES,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BIGBY,CHARLES
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: CHATFIELD

DATE/TIME OF TEST: 11/21/2019 02:39:57

Diagnostics Check		OK	02:33
Air Blank		0.000	02:34
Control Test		0.080	02:34
Air Blank		0.000	02:34
Subject Sample	# 1	0.152	02:35
Breath Volume		2.691	02:35
Air Blank		0.000	02:36
Air Blank		0.000	02:37
Subject Sample	# 2	0.154	02:38
Breath Volume		3.148	02:38
Air Blank		0.000	02:39
Control Test		0.080	02:39
Air Blank		0.000	02:39
Diagnostics Check		OK	02:39

Data Download Date/Time: 12/17/2019 09:15

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 11/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: GODSHALL

DATE/TIME OF TEST: 11/21/2019 04:15:21

Diagnostics Check		OK	04:07	
Air Blank		0.000	04:07	
Control Test		0.080	04:07	
Air Blank		0.000	04:08	
Subject Sample	# 1	0.054	04:09	
Breath Volume		2.351	04:09	
Air Blank		0.000	04:09	
Air Blank		0.000	04:11	
Subject Sample	# 2		04:13	Subject Test Refused
Breath Volume		0.000	04:13	
Air Blank		0.000	04:14	
Control Test		0.080	04:14	
Air Blank		0.000	04:15	
Diagnostics Check		OK	04:15	

Data Download Date/Time: 12/17/2019 09:15

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: LEFEVRE

DATE/TIME OF TEST: 11/22/2019 01:56:12

Diagnostics Check		OK	01:49
Air Blank		0.000	01:50
Control Test		0.080	01:50
Air Blank		0.000	01:51
Subject Sample	# 1	0.196	01:51
Breath Volume		2.546	01:51
Air Blank		0.000	01:52
Air Blank		0.000	01:54
Subject Sample	# 2	0.198	01:54
Breath Volume		1.972	01:54
Air Blank		0.000	01:55
Control Test		0.078	01:55
Air Blank		0.000	01:56
Diagnostics Check		OK	01:56

Data Download Date/Time: 12/17/2019 09:15

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGBY,CHARLES,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: CUMMINGS III,GARY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 11/22/2019 04:00:31

Diagnostics Check		OK	03:52
Air Blank		0.000	03:52
Control Test		0.080	03:53
Air Blank		0.000	03:53
Subject Sample	# 1	0.125	03:54
Breath Volume		1.515	03:54
Air Blank		0.000	03:55
Air Blank		0.000	03:56
Subject Sample	# 2	0.126	03:58
Breath Volume		2.351	03:58
Air Blank		0.000	03:59
Control Test		0.080	03:59
Air Blank		0.000	04:00
Diagnostics Check		OK	04:00

Data Download Date/Time: 12/17/2019 09:15

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 11/23/2019 01:12:05

Diagnostics Check		OK	01:05
Air Blank		0.000	01:06
Control Test		0.080	01:06
Air Blank		0.000	01:07
Subject Sample	# 1	0.166	01:07
Breath Volume		3.171	01:07
Air Blank		0.000	01:08
Air Blank		0.000	01:10
Subject Sample	# 2	0.164	01:10
Breath Volume		3.054	01:10
Air Blank		0.000	01:11
Control Test		0.079	01:11
Air Blank		0.000	01:11
Diagnostics Check		OK	01:12

Data Download Date/Time: 12/17/2019 09:15

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 19:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: PINA

DATE/TIME OF TEST: 11/24/2019 20:56:19

Diagnostics Check		OK	20:46
Air Blank		0.000	20:46
Control Test		0.080	20:46
Air Blank		0.000	20:47
Subject Sample	# 1	0.255	20:47
Breath Volume		1.460	20:47
Air Blank		0.000	20:48
Air Blank		0.000	20:50
Subject Sample	# 2	0.225	20:51
Breath Volume		1.640	20:51
Air Blank		0.000	20:51
Air Blank		0.000	20:53
Subject Sample	# 3	0.233	20:54
Breath Volume		2.613	20:54
Air Blank		0.000	20:55
Control Test		0.080	20:55
Air Blank		0.000	20:56
Diagnostics Check		OK	20:56

Data Download Date/Time: 12/17/2019 09:16

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAAVILAINEN,JOSEPH,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: PAAVILAINEN,JOSEPH
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: SIPLIN

DATE/TIME OF TEST: 11/24/2019 23:07:12

Diagnostics Check		OK	23:00
Air Blank		0.000	23:00
Control Test		0.080	23:01
Air Blank		0.000	23:01
Subject Sample	# 1	0.196	23:02
Breath Volume		3.246	23:02
Air Blank		0.000	23:03
Air Blank		0.000	23:05
Subject Sample	# 2	0.196	23:05
Breath Volume		2.437	23:05
Air Blank		0.000	23:06
Control Test		0.079	23:06
Air Blank		0.000	23:07
Diagnostics Check		OK	23:07

Data Download Date/Time: 12/17/2019 09:16

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 11/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: WHITAKER

DATE/TIME OF TEST: 11/27/2019 04:22:31

Diagnostics Check		OK	04:15
Air Blank		0.000	04:16
Control Test		0.080	04:16
Air Blank		0.000	04:17
Subject Sample	# 1	0.179	04:18
Breath Volume		2.597	04:18
Air Blank		0.000	04:18
Air Blank		0.000	04:20
Subject Sample	# 2	0.181	04:20
Breath Volume		2.714	04:20
Air Blank		0.000	04:21
Control Test		0.080	04:21
Air Blank		0.000	04:22
Diagnostics Check		OK	04:22

Data Download Date/Time: 12/17/2019 09:16

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,CODY,R
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BARRETT,CODY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: COTHERN

DATE/TIME OF TEST: 11/27/2019 23:42:23

Diagnostics Check		OK	23:36
Air Blank		0.000	23:36
Control Test		0.080	23:36
Air Blank		0.000	23:37
Subject Sample	# 1	0.176	23:37
Breath Volume		3.300	23:37
Air Blank		0.000	23:38
Air Blank		0.000	23:40
Subject Sample	# 2	0.182	23:40
Breath Volume		3.234	23:40
Air Blank		0.000	23:41
Control Test		0.079	23:41
Air Blank		0.000	23:42
Diagnostics Check		OK	23:42

Data Download Date/Time: 12/17/2019 09:16

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGBY,CHARLES,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BIGBY,CHARLES
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 11/29/2019 23:19:49

Diagnostics Check		OK	23:16	
Air Blank		0.000	23:17	
Control Test		0.081	23:17	
Air Blank		0.000	23:18	
Subject Sample	# 1		23:18	Subject Test Refused
Breath Volume		0.000	23:18	
Air Blank		0.000	23:18	
Control Test		0.080	23:19	
Air Blank		0.000	23:19	
Diagnostics Check		OK	23:19	

Data Download Date/Time: 12/17/2019 09:16

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 11/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: RAMOS RAMIREZ

DATE/TIME OF TEST: 11/30/2019 00:34:53

Diagnostics Check		OK	00:28
Air Blank		0.000	00:28
Control Test		0.080	00:29
Air Blank		0.000	00:29
Subject Sample	# 1	0.309	00:30
Breath Volume		2.304	00:30
Air Blank		0.000	00:31
Air Blank		0.000	00:32
Subject Sample	# 2	0.326	00:33
Breath Volume		2.121	00:33
Air Blank		0.000	00:33
Control Test		0.080	00:34
Air Blank		0.000	00:34
Diagnostics Check		OK	00:34

Data Download Date/Time: 12/17/2019 09:16

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGBY,CHARLES,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BIGBY,CHARLES
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: MCLANE

DATE/TIME OF TEST: 11/30/2019 02:13:36

Diagnostics Check		OK	02:10	
Air Blank		0.000	02:11	
Control Test		0.080	02:11	
Air Blank		0.000	02:11	
Subject Sample	# 1		02:12	Subject Test Refused
Breath Volume		0.000	02:12	
Air Blank		0.000	02:12	
Control Test		0.080	02:13	
Air Blank		0.000	02:13	
Diagnostics Check		OK	02:13	

Data Download Date/Time: 12/17/2019 09:16

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUMMINGS,GARY,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: CUMMINGS,GARY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 06:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:12

SUBJECT'S LAST NAME: CARSON

DATE/TIME OF TEST: 11/30/2019 07:36:32

Diagnostics Check		OK	07:33	
Air Blank		0.000	07:33	
Control Test		0.080	07:34	
Air Blank		0.000	07:34	
Subject Sample	# 1	0.201	07:35	
Breath Volume		2.199	07:35	
Air Blank			07:36	Purge Fail
Air Blank		0.000	07:36	

Data Download Date/Time: 12/17/2019 09:16

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUMMINGS,GARY,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: CUMMINGS,GARY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 06:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:13

SUBJECT'S LAST NAME: CARSON

DATE/TIME OF TEST: 11/30/2019 07:45:21

Diagnostics Check		OK	07:38
Air Blank		0.000	07:39
Control Test		0.079	07:39
Air Blank		0.000	07:40
Subject Sample	# 1	0.189	07:40
Breath Volume		2.738	07:40
Air Blank		0.000	07:41
Air Blank		0.000	07:43
Subject Sample	# 2	0.195	07:43
Breath Volume		3.722	07:43
Air Blank		0.000	07:44
Control Test		0.077	07:44
Air Blank		0.000	07:45
Diagnostics Check		OK	07:45

Data Download Date/Time: 12/17/2019 09:16

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGBY,CHARLES,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BIGBY,CHARLES
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: CRUZ-LOPEZ

DATE/TIME OF TEST: 11/30/2019 23:17:18

Diagnostics Check		OK	23:10
Air Blank		0.000	23:10
Control Test		0.080	23:11
Air Blank		0.000	23:11
Subject Sample	# 1	0.268	23:12
Breath Volume		2.191	23:12
Air Blank		0.000	23:13
Air Blank		0.000	23:15
Subject Sample	# 2	0.269	23:15
Breath Volume		2.253	23:15
Air Blank		0.000	23:16
Control Test		0.080	23:16
Air Blank		0.000	23:17
Diagnostics Check		OK	23:17

Data Download Date/Time: 12/17/2019 09:16

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGBY,CHARLES,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BIGBY,CHARLES
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: HERNANDEZ HERNANDEZ

DATE/TIME OF TEST: 11/02/2019 02:51:31

Diagnostics Check		OK	02:43
Air Blank		0.000	02:44
Control Test		0.081	02:44
Air Blank		0.000	02:45
Subject Sample	# 1	0.229	02:46
Breath Volume		1.625	02:46
Air Blank		0.000	02:47
Air Blank		0.000	02:48
Subject Sample	# 2	0.235	02:49
Breath Volume		1.593	02:49
Air Blank		0.000	02:50
Control Test		0.079	02:50
Air Blank		0.000	02:51
Diagnostics Check		OK	02:51

Data Download Date/Time: 11/18/2019 12:58

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KING,GRACE,N
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: KING,GRACE
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: ROSENBLATT

DATE/TIME OF TEST: 11/02/2019 03:09:22

Diagnostics Check		OK	03:06	
Air Blank		0.000	03:06	
Control Test		0.079	03:06	
Air Blank		0.000	03:07	
Subject Sample	# 1		03:07	Subject Test Refused
Breath Volume		0.000	03:07	
Air Blank		0.000	03:08	
Control Test		0.081	03:08	
Air Blank		0.000	03:09	
Diagnostics Check		OK	03:09	

Data Download Date/Time: 11/18/2019 12:58

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGBY,CHARLES,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BIGBY,CHARLES
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: AVERY

DATE/TIME OF TEST: 11/03/2019 01:49:06

Diagnostics Check		OK	01:42
Air Blank		0.000	01:43
Control Test		0.081	01:43
Air Blank		0.000	01:44
Subject Sample	# 1	0.230	01:44
Breath Volume		2.707	01:44
Air Blank		0.000	01:45
Air Blank		0.000	01:46
Subject Sample	# 2	0.235	01:47
Breath Volume		2.832	01:47
Air Blank		0.000	01:48
Control Test		0.080	01:48
Air Blank		0.000	01:49
Diagnostics Check		OK	01:49

Data Download Date/Time: 11/18/2019 12:58

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KING,GRACE,N
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: KING,GRACE
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: CRAVENS

DATE/TIME OF TEST: 11/03/2019 01:07:20

Diagnostics Check		OK	01:00
Air Blank		0.000	01:01
Control Test		0.080	01:01
Air Blank		0.000	01:01
Subject Sample	# 1	0.062	01:02
Breath Volume		1.542	01:02
Air Blank		0.000	01:03
Air Blank		0.000	01:05
Subject Sample	# 2	0.068	01:05
Breath Volume		2.882	01:05
Air Blank		0.000	01:06
Control Test		0.078	01:06
Air Blank		0.000	01:07
Diagnostics Check		OK	01:07

Data Download Date/Time: 11/18/2019 12:58

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: LOWE

DATE/TIME OF TEST: 11/03/2019 01:44:05

Diagnostics Check		OK	01:37
Air Blank		0.000	01:37
Control Test		0.080	01:37
Air Blank		0.000	01:38
Subject Sample	# 1	0.144	01:38
Breath Volume		1.242	01:38
Air Blank		0.000	01:39
Air Blank		0.000	01:41
Subject Sample	# 2	0.155	01:42
Breath Volume		1.128	01:42
Air Blank		0.000	01:43
Control Test		0.080	01:43
Air Blank		0.000	01:44
Diagnostics Check		OK	01:44

Data Download Date/Time: 11/18/2019 12:58

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 20:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:49

SUBJECT'S LAST NAME: PEREZ PEREZ

DATE/TIME OF TEST: 11/03/2019 21:17:36

Diagnostics Check		OK	21:09
Air Blank		0.000	21:09
Control Test		0.081	21:10
Air Blank		0.000	21:10
Subject Sample	# 1	0.219	21:11
Breath Volume		1.847	21:11
Air Blank		0.000	21:12
Air Blank		0.000	21:14
Subject Sample	# 2	0.216	21:15
Breath Volume		2.312	21:15
Air Blank		0.000	21:16
Control Test		0.079	21:16
Air Blank		0.000	21:17
Diagnostics Check		OK	21:17

Data Download Date/Time: 11/18/2019 12:58

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 20:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:49

SUBJECT'S LAST NAME: HERRERA REYES

DATE/TIME OF TEST: 11/03/2019 21:25:18

Diagnostics Check		OK	21:18
Air Blank		0.000	21:19
Control Test		0.078	21:19
Air Blank		0.000	21:20
Subject Sample	# 1	0.069	21:20
Breath Volume		3.210	21:20
Air Blank		0.000	21:21
Air Blank		0.000	21:23
Subject Sample	# 2	0.068	21:23
Breath Volume		3.695	21:23
Air Blank		0.000	21:24
Control Test		0.080	21:24
Air Blank		0.000	21:25
Diagnostics Check		OK	21:25

Data Download Date/Time: 11/18/2019 12:58

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: DORION

DATE/TIME OF TEST: 11/04/2019 04:10:51

Diagnostics Check		OK	04:04
Air Blank		0.000	04:04
Control Test		0.081	04:05
Air Blank		0.000	04:05
Subject Sample	# 1	0.134	04:06
Breath Volume		2.058	04:06
Air Blank		0.000	04:07
Air Blank		0.000	04:08
Subject Sample	# 2	0.134	04:09
Breath Volume		2.007	04:09
Air Blank		0.000	04:09
Control Test		0.080	04:10
Air Blank		0.000	04:10
Diagnostics Check		OK	04:10

Data Download Date/Time: 11/18/2019 12:58

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER III,RICHARD,F
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MOORE,ADAM
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 10/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 11/06/2019 00:39:34

Diagnostics Check		OK	00:36	
Air Blank		0.000	00:37	
Control Test		0.081	00:37	
Air Blank		0.000	00:37	
Subject Sample	# 1		00:38	Subject Test Refused
Breath Volume		0.000	00:38	
Air Blank		0.000	00:38	
Control Test		0.081	00:38	
Air Blank		0.000	00:39	
Diagnostics Check		OK	00:39	

Data Download Date/Time: 11/18/2019 12:58

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 17:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: PLUFF

DATE/TIME OF TEST: 11/06/2019 18:49:21

Diagnostics Check		OK	18:34	
Air Blank		0.000	18:34	
Control Test		0.081	18:35	
Air Blank		0.000	18:35	
Subject Sample	# 1	0.217	18:36	
Breath Volume		1.460	18:36	
Air Blank		0.000	18:36	
Air Blank		0.000	18:38	
Subject Sample	# 2	0.197	18:41	Volume Not Met
Breath Volume		1.148	18:41	
Air Blank		0.000	18:42	
Air Blank		0.000	18:44	
Subject Sample	# 3	0.170	18:47	Volume Not Met
Breath Volume		0.195	18:47	
Air Blank		0.000	18:48	
Control Test		0.078	18:48	
Air Blank		0.000	18:49	
Diagnostics Check		OK	18:49	

Data Download Date/Time: 11/18/2019 12:58

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 17:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: PLUFF

DATE/TIME OF TEST: 11/06/2019 18:58:17

Diagnostics Check		OK	18:51
Air Blank		0.000	18:51
Control Test		0.079	18:51
Air Blank		0.000	18:52
Subject Sample	# 1	0.180	18:52
Breath Volume		1.828	18:52
Air Blank		0.000	18:53
Air Blank		0.000	18:55
Subject Sample	# 2	0.199	18:56
Breath Volume		1.886	18:56
Air Blank		0.000	18:57
Control Test		0.079	18:57
Air Blank		0.000	18:58
Diagnostics Check		OK	18:58

Data Download Date/Time: 11/18/2019 12:58

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENYAN,M,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MASON,S
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 10/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: BURNHAM

DATE/TIME OF TEST: 11/08/2019 02:00:41

Diagnostics Check		OK	01:53
Air Blank		0.000	01:53
Control Test		0.081	01:54
Air Blank		0.000	01:54
Subject Sample	# 1	0.018	01:55
Breath Volume		2.375	01:55
Air Blank		0.000	01:56
Air Blank		0.000	01:58
Subject Sample	# 2	0.017	01:59
Breath Volume		2.031	01:59
Air Blank		0.000	01:59
Control Test		0.080	02:00
Air Blank		0.000	02:00
Diagnostics Check		OK	02:00

Data Download Date/Time: 11/18/2019 12:58

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,CODY,R
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BARRETT,CODY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: COLES

DATE/TIME OF TEST: 11/10/2019 01:28:04

Diagnostics Check		OK	01:21
Air Blank		0.000	01:21
Control Test		0.081	01:22
Air Blank		0.000	01:22
Subject Sample	# 1	0.050	01:23
Breath Volume		2.464	01:23
Air Blank		0.000	01:23
Air Blank		0.000	01:25
Subject Sample	# 2	0.047	01:26
Breath Volume		1.433	01:26
Air Blank		0.000	01:27
Control Test		0.079	01:27
Air Blank		0.000	01:27
Diagnostics Check		OK	01:28

Data Download Date/Time: 11/18/2019 12:58

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: KLOBUCHAR,RICHARD
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 10/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: GARRINGER

DATE/TIME OF TEST: 11/12/2019 03:22:44

Diagnostics Check		OK	03:15
Air Blank		0.000	03:16
Control Test		0.081	03:16
Air Blank		0.000	03:17
Subject Sample	# 1	0.103	03:17
Breath Volume		2.445	03:17
Air Blank		0.000	03:18
Air Blank		0.000	03:20
Subject Sample	# 2	0.102	03:21
Breath Volume		1.906	03:21
Air Blank		0.000	03:21
Control Test		0.080	03:22
Air Blank		0.000	03:22
Diagnostics Check		OK	03:22

Data Download Date/Time: 11/18/2019 12:58

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: GRAY

DATE/TIME OF TEST: 11/14/2019 02:58:38

Diagnostics Check		OK	02:52
Air Blank		0.000	02:52
Control Test		0.081	02:52
Air Blank		0.000	02:53
Subject Sample	# 1	0.173	02:53
Breath Volume		2.656	02:53
Air Blank		0.000	02:54
Air Blank		0.000	02:56
Subject Sample	# 2	0.168	02:56
Breath Volume		3.640	02:56
Air Blank		0.000	02:57
Control Test		0.080	02:58
Air Blank		0.000	02:58
Diagnostics Check		OK	02:58

Data Download Date/Time: 11/18/2019 12:58

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGBY,CHARLES,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BIGBY,CHARLES
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: SALE

DATE/TIME OF TEST: 11/16/2019 02:03:13

Diagnostics Check		OK	01:56
Air Blank		0.000	01:56
Control Test		0.080	01:57
Air Blank		0.000	01:57
Subject Sample	# 1	0.193	01:58
Breath Volume		1.562	01:58
Air Blank		0.000	01:59
Air Blank		0.000	02:01
Subject Sample	# 2	0.182	02:01
Breath Volume		1.429	02:01
Air Blank		0.000	02:02
Control Test		0.080	02:02
Air Blank		0.000	02:03
Diagnostics Check		OK	02:03

Data Download Date/Time: 11/18/2019 12:58

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 11/16/2019 04:19:02

Diagnostics Check		OK	04:11
Air Blank		0.000	04:12
Control Test		0.080	04:12
Air Blank		0.000	04:12
Subject Sample	# 1	0.210	04:13
Breath Volume		1.156	04:13
Air Blank		0.000	04:14
Air Blank		0.000	04:16
Subject Sample	# 2	0.203	04:17
Breath Volume		1.578	04:17
Air Blank		0.000	04:18
Control Test		0.079	04:18
Air Blank		0.000	04:18
Diagnostics Check		OK	04:19

Data Download Date/Time: 11/18/2019 12:58

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGBY,CHARLES,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: CUMMINGSIII,GARY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: BALLARD

DATE/TIME OF TEST: 11/16/2019 04:34:16

Diagnostics Check		OK	04:27
Air Blank		0.000	04:28
Control Test		0.080	04:28
Air Blank		0.000	04:28
Subject Sample	# 1	0.193	04:29
Breath Volume		2.445	04:29
Air Blank		0.000	04:30
Air Blank		0.000	04:31
Subject Sample	# 2	0.201	04:32
Breath Volume		2.003	04:32
Air Blank		0.000	04:33
Control Test		0.079	04:33
Air Blank		0.000	04:34
Diagnostics Check		OK	04:34

Data Download Date/Time: 11/18/2019 12:58

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 19:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: AGUIRRE AVELLANEDA

DATE/TIME OF TEST: 11/16/2019 20:27:45

Diagnostics Check		OK	20:20	
Air Blank		0.000	20:21	
Control Test		0.081	20:21	
Air Blank		0.000	20:22	
Subject Sample	# 1	0.081	20:23	
Breath Volume		1.398	20:23	
Air Blank		0.000	20:23	
Air Blank		0.000	20:25	
Subject Sample	# 2		20:26	Slope Not Met
Breath Volume		0.996	20:26	
Air Blank		0.000	20:26	
Control Test		0.081	20:27	
Air Blank		0.000	20:27	
Diagnostics Check		OK	20:27	

Data Download Date/Time: 11/18/2019 12:58

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 19:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: AGUIRRE AVELLANEDA

DATE/TIME OF TEST: 11/16/2019 20:35:11

Diagnostics Check		OK	20:29
Air Blank		0.000	20:29
Control Test		0.080	20:29
Air Blank		0.000	20:30
Subject Sample	# 1	0.071	20:30
Breath Volume		1.488	20:30
Air Blank		0.000	20:31
Air Blank		0.000	20:33
Subject Sample	# 2	0.070	20:33
Breath Volume		2.457	20:33
Air Blank		0.000	20:34
Control Test		0.080	20:34
Air Blank		0.000	20:35
Diagnostics Check		OK	20:35

Data Download Date/Time: 11/18/2019 12:58

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGBY,CHARLES,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BIGBY,CHARLES
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: VALERIO-BETANCOURT

DATE/TIME OF TEST: 11/17/2019 00:01:17

Diagnostics Check		OK	23:58	
Air Blank		0.000	23:58	
Control Test		0.081	23:58	
Air Blank		0.000	23:59	
Subject Sample	# 1		23:59	Subject Test Refused
Breath Volume		0.000	23:59	
Air Blank		0.000	00:00	
Control Test		0.081	00:00	
Air Blank		0.000	00:01	
Diagnostics Check		OK	00:01	

Data Download Date/Time: 11/18/2019 12:58

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: RUIZ AYALA

DATE/TIME OF TEST: 11/17/2019 02:21:45

Diagnostics Check		OK	02:09	
Air Blank		0.000	02:10	
Control Test		0.081	02:10	
Air Blank		0.000	02:11	
Subject Sample	# 1	0.122	02:14	Volume Not Met
Breath Volume		0.015	02:14	
Air Blank		0.000	02:15	
Air Blank		0.000	02:16	
Subject Sample	# 2	0.147	02:20	Volume Not Met
Breath Volume		0.558	02:20	
Air Blank		0.000	02:20	
Control Test		0.080	02:21	
Air Blank		0.000	02:21	
Diagnostics Check		OK	02:21	

Data Download Date/Time: 11/18/2019 12:58

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: RUIZ AYALA

DATE/TIME OF TEST: 11/17/2019 02:32:32

Diagnostics Check		OK	02:22
Air Blank		0.000	02:23
Control Test		0.079	02:23
Air Blank		0.000	02:24
Subject Sample	# 1	0.146	02:25
Breath Volume		1.210	02:25
Air Blank		0.000	02:26
Air Blank		0.000	02:27
Subject Sample	# 2	0.139	02:30
Breath Volume		1.218	02:30
Air Blank		0.000	02:31
Control Test		0.080	02:31
Air Blank		0.000	02:32
Diagnostics Check		OK	02:32

Data Download Date/Time: 11/18/2019 12:58

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGBY,CHARLES,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BIGBY,CHARLES
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 11/17/2019 03:32:50

Diagnostics Check		OK	03:26
Air Blank		0.000	03:27
Control Test		0.080	03:27
Air Blank		0.000	03:27
Subject Sample	# 1	0.163	03:28
Breath Volume		2.777	03:28
Air Blank		0.000	03:29
Air Blank		0.000	03:30
Subject Sample	# 2	0.161	03:31
Breath Volume		2.929	03:31
Air Blank		0.000	03:31
Control Test		0.081	03:32
Air Blank		0.000	03:32
Diagnostics Check		OK	03:32

Data Download Date/Time: 11/18/2019 12:58

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 10/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: HERNANDEZ LOPEZ

DATE/TIME OF TEST: 11/17/2019 03:51:30

Diagnostics Check		OK	03:44
Air Blank		0.000	03:44
Control Test		0.080	03:45
Air Blank		0.000	03:45
Subject Sample	# 1	0.128	03:46
Breath Volume		1.519	03:46
Air Blank		0.000	03:47
Air Blank		0.000	03:49
Subject Sample	# 2	0.128	03:49
Breath Volume		2.000	03:49
Air Blank		0.000	03:50
Control Test		0.080	03:50
Air Blank		0.000	03:51
Diagnostics Check		OK	03:51

Data Download Date/Time: 11/18/2019 12:58

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENYAN,M,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: KENYAN,MA
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 05:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:45

SUBJECT'S LAST NAME: VAUGHAN

DATE/TIME OF TEST: 11/17/2019 06:12:44

Diagnostics Check		OK	06:05
Air Blank		0.000	06:06
Control Test		0.081	06:06
Air Blank		0.000	06:07
Subject Sample	# 1	0.238	06:07
Breath Volume		2.031	06:07
Air Blank		0.000	06:08
Air Blank		0.000	06:10
Subject Sample	# 2	0.224	06:10
Breath Volume		1.460	06:10
Air Blank		0.000	06:11
Control Test		0.080	06:12
Air Blank		0.000	06:12
Diagnostics Check		OK	06:12

Data Download Date/Time: 11/18/2019 12:58

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLER,SANDRA,J
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: CARDUCCI,TERES
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 08:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:24

SUBJECT'S LAST NAME: VASQUEZ

DATE/TIME OF TEST: 11/17/2019 11:05:52

Diagnostics Check		OK	11:02	
Air Blank		0.000	11:03	
Control Test		0.081	11:03	
Air Blank		0.000	11:03	
Subject Sample	# 1		11:05	Sequence Aborted
Breath Volume		0.000	11:05	
Air Blank		0.000	11:05	

Data Download Date/Time: 11/18/2019 12:58

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGBY,CHARLES,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BIGBY,CHARLES
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 21:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:14

SUBJECT'S LAST NAME: BURKE

DATE/TIME OF TEST: 11/17/2019 22:00:40

Diagnostics Check		OK	21:47	
Air Blank		0.000	21:48	
Control Test		0.080	21:48	
Air Blank		0.000	21:48	
Subject Sample	# 1	0.102	21:52	Volume Not Met
Breath Volume		0.187	21:52	
Air Blank		0.000	21:52	
Air Blank		0.000	21:54	
Subject Sample	# 2	0.150	21:55	
Breath Volume		1.285	21:55	
Air Blank		0.000	21:56	
Air Blank		0.000	21:58	
Subject Sample	# 3	0.150	21:59	
Breath Volume		1.710	21:59	
Air Blank		0.000	21:59	
Control Test		0.080	22:00	
Air Blank		0.000	22:00	
Diagnostics Check		OK	22:00	

Data Download Date/Time: 11/18/2019 12:58

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 10/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: RACHIDE

DATE/TIME OF TEST: 11/18/2019 02:30:54

Diagnostics Check		OK	02:23
Air Blank		0.000	02:23
Control Test		0.080	02:24
Air Blank		0.000	02:24
Subject Sample	# 1	0.288	02:25
Breath Volume		2.554	02:25
Air Blank		0.000	02:26
Air Blank		0.000	02:28
Subject Sample	# 2	0.278	02:29
Breath Volume		2.628	02:29
Air Blank		0.000	02:29
Control Test		0.080	02:30
Air Blank		0.000	02:30
Diagnostics Check		OK	02:30

Data Download Date/Time: 11/18/2019 12:58

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: EKE JR

DATE/TIME OF TEST: 11/01/2019 02:38:17

Diagnostics Check		OK	02:30
Air Blank		0.000	02:31
Control Test		0.080	02:31
Air Blank		0.000	02:32
Subject Sample	# 1	0.147	02:32
Breath Volume		1.289	02:32
Air Blank		0.000	02:33
Air Blank		0.000	02:35
Subject Sample	# 2	0.155	02:36
Breath Volume		2.394	02:36
Air Blank		0.000	02:37
Control Test		0.080	02:37
Air Blank		0.000	02:38
Diagnostics Check		OK	02:38

Data Download Date/Time: 11/06/2019 18:31

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 11/02/2019 00:44:26

Diagnostics Check		OK	00:37
Air Blank		0.000	00:37
Control Test		0.080	00:38
Air Blank		0.000	00:38
Subject Sample	# 1	0.158	00:39
Breath Volume		1.960	00:39
Air Blank		0.000	00:40
Air Blank		0.000	00:42
Subject Sample	# 2	0.167	00:42
Breath Volume		3.359	00:42
Air Blank		0.000	00:43
Control Test		0.080	00:43
Air Blank		0.000	00:44
Diagnostics Check		OK	00:44

Data Download Date/Time: 11/06/2019 18:31

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TIBURCIO,I
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: CHRISTIE

DATE/TIME OF TEST: 11/02/2019 01:44:41

Diagnostics Check		OK	01:38
Air Blank		0.000	01:38
Control Test		0.080	01:39
Air Blank		0.000	01:39
Subject Sample	# 1	0.244	01:40
Breath Volume		2.855	01:40
Air Blank		0.000	01:40
Air Blank		0.000	01:42
Subject Sample	# 2	0.251	01:43
Breath Volume		2.460	01:43
Air Blank		0.000	01:43
Control Test		0.079	01:44
Air Blank		0.000	01:44
Diagnostics Check		OK	01:44

Data Download Date/Time: 11/06/2019 18:31

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ROTHCHILD,F
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: STORM

DATE/TIME OF TEST: 11/02/2019 03:51:35

Diagnostics Check		OK	03:43
Air Blank		0.000	03:44
Control Test		0.080	03:44
Air Blank		0.000	03:45
Subject Sample	# 1	0.066	03:47
Breath Volume		1.652	03:47
Air Blank		0.000	03:47
Air Blank		0.000	03:49
Subject Sample	# 2	0.066	03:50
Breath Volume		2.800	03:50
Air Blank		0.000	03:50
Control Test		0.080	03:50
Air Blank		0.000	03:51
Diagnostics Check		OK	03:51

Data Download Date/Time: 11/06/2019 18:31

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ARCHAMBEAU,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: SPISSINGER

DATE/TIME OF TEST: 11/02/2019 04:53:27

Diagnostics Check		OK	04:45
Air Blank		0.000	04:46
Control Test		0.080	04:46
Air Blank		0.000	04:47
Subject Sample	# 1	0.117	04:47
Breath Volume		1.437	04:47
Air Blank		0.000	04:48
Air Blank		0.000	04:50
Subject Sample	# 2	0.119	04:51
Breath Volume		1.300	04:51
Air Blank		0.000	04:52
Control Test		0.080	04:52
Air Blank		0.000	04:53
Diagnostics Check		OK	04:53

Data Download Date/Time: 11/06/2019 18:31

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: SHAW

DATE/TIME OF TEST: 11/02/2019 23:47:44

Diagnostics Check		OK	23:40
Air Blank		0.000	23:40
Control Test		0.080	23:41
Air Blank		0.000	23:41
Subject Sample	# 1	0.009	23:43
Breath Volume		1.894	23:43
Air Blank		0.000	23:43
Air Blank		0.000	23:45
Subject Sample	# 2	0.000	23:46
Breath Volume		1.109	23:46
Air Blank		0.000	23:46
Control Test		0.080	23:47
Air Blank		0.000	23:47
Diagnostics Check		OK	23:47

Data Download Date/Time: 11/06/2019 18:32

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES AVERY,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: MELO SANTANA

DATE/TIME OF TEST: 11/03/2019 03:44:02

Diagnostics Check		OK	03:37
Air Blank		0.000	03:37
Control Test		0.080	03:38
Air Blank		0.000	03:38
Subject Sample	# 1	0.113	03:39
Breath Volume		1.492	03:39
Air Blank		0.000	03:39
Air Blank		0.000	03:41
Subject Sample	# 2	0.117	03:42
Breath Volume		1.617	03:42
Air Blank		0.000	03:43
Control Test		0.080	03:43
Air Blank		0.000	03:43
Diagnostics Check		OK	03:44

Data Download Date/Time: 11/06/2019 18:32

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES AVERY,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VANTREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: NOAH ABANA

DATE/TIME OF TEST: 11/04/2019 00:27:46

Diagnostics Check		OK	00:21
Air Blank		0.000	00:21
Control Test		0.080	00:22
Air Blank		0.000	00:22
Subject Sample	# 1	0.257	00:23
Breath Volume		1.210	00:23
Air Blank		0.000	00:24
Air Blank		0.000	00:25
Subject Sample	# 2	0.257	00:26
Breath Volume		1.597	00:26
Air Blank		0.000	00:26
Control Test		0.079	00:27
Air Blank		0.000	00:27
Diagnostics Check		OK	00:27

Data Download Date/Time: 11/06/2019 18:32

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,CARL,L
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: PARKER,CARL
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 11/17/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:26

SUBJECT'S LAST NAME: HARDEE

DATE/TIME OF TEST: 11/20/2019 22:31:22

Diagnostics Check		OK	22:24
Air Blank		0.000	22:24
Control Test		0.080	22:24
Air Blank		0.000	22:25
Subject Sample	# 1	0.127	22:26
Breath Volume		1.179	22:26
Air Blank		0.000	22:27
Air Blank		0.000	22:28
Subject Sample	# 2	0.125	22:29
Breath Volume		1.437	22:29
Air Blank		0.000	22:30
Control Test		0.079	22:30
Air Blank		0.000	22:31
Diagnostics Check		OK	22:31

Data Download Date/Time: 12/31/2019 19:20

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONG,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: CAPELA,JAYD
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 11/17/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 12:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:33

SUBJECT'S LAST NAME: MONAHAN

DATE/TIME OF TEST: 11/21/2019 14:02:04

Diagnostics Check		OK	13:56	
Air Blank		0.000	13:56	
Control Test		0.080	13:57	
Air Blank		0.000	13:57	
Subject Sample	# 1	0.193	14:00	Volume Not Met
Breath Volume		0.281	14:00	
Air Blank			14:01	Purge Fail
Air Blank		0.000	14:02	

Data Download Date/Time: 12/31/2019 19:20

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONG,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: CAPELA,JAYD
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 11/17/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 12:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:33

SUBJECT'S LAST NAME: MONAHAN

DATE/TIME OF TEST: 11/21/2019 14:17:11

Diagnostics Check		OK	14:09	
Air Blank		0.000	14:09	
Control Test		0.077	14:09	
Air Blank		0.000	14:10	
Subject Sample	# 1	0.272	14:11	
Breath Volume		1.492	14:11	
Air Blank		0.000	14:12	
Air Blank		0.000	14:14	
Subject Sample	# 2	0.260	14:15	
Breath Volume		1.644	14:15	
Air Blank		0.000	14:16	
Control Test		0.074	14:16	Control Outside Tolerance
Air Blank		0.000	14:17	

Data Download Date/Time: 12/31/2019 19:20

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONG,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: CAPELA,JAYD
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 11/17/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 12:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:33

SUBJECT'S LAST NAME: MONAHAN

DATE/TIME OF TEST: 11/21/2019 14:37:56

Diagnostics Check		OK	14:24	
Air Blank		0.000	14:24	
Control Test		0.078	14:25	
Air Blank		0.000	14:25	
Subject Sample	# 1	0.272	14:26	
Breath Volume		1.644	14:26	
Air Blank		0.000	14:27	
Air Blank		0.000	14:29	
Subject Sample	# 2	0.204	14:32	Volume Not Met
Breath Volume		0.753	14:32	
Air Blank		0.000	14:33	
Air Blank		0.000	14:35	
Subject Sample	# 3	0.252	14:36	
Breath Volume		1.687	14:36	
Air Blank			14:37	Purge Fail
Air Blank		0.000	14:37	

Data Download Date/Time: 12/31/2019 19:20

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONG,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: CAPELA,JAYD
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 11/17/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 12:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:33

SUBJECT'S LAST NAME: MONAHAN

DATE/TIME OF TEST: 11/21/2019 14:48:42

Diagnostics Check		OK	14:42	
Air Blank		0.000	14:43	
Control Test		0.075	14:43	
Air Blank		0.000	14:44	
Subject Sample	# 1	0.201	14:47	Volume Not Met
Breath Volume		0.593	14:47	
Air Blank			14:48	Purge Fail
Air Blank		0.000	14:48	

Data Download Date/Time: 12/31/2019 19:20

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NICOLAS,MITA,
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: SPINA,TYLOR
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 11/17/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: MEZA

DATE/TIME OF TEST: 11/22/2019 21:39:05

Diagnostics Check		OK	21:36	
Air Blank		0.000	21:36	
Control Test		0.079	21:36	
Air Blank		0.000	21:37	
Subject Sample	# 1		21:37	Subject Test Refused
Breath Volume		0.000	21:37	
Air Blank		0.000	21:38	
Control Test		0.079	21:38	
Air Blank		0.000	21:38	
Diagnostics Check		OK	21:39	

Data Download Date/Time: 12/31/2019 19:20

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TWING,BRANDON,R
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: CONWAY,P
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 11/17/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: HAWKESWORTH

DATE/TIME OF TEST: 11/23/2019 04:38:29

Diagnostics Check		OK	04:29
Air Blank		0.000	04:30
Control Test		0.079	04:30
Air Blank		0.000	04:31
Subject Sample	# 1	0.152	04:33
Breath Volume		3.328	04:33
Air Blank		0.000	04:34
Air Blank		0.000	04:36
Subject Sample	# 2	0.139	04:36
Breath Volume		2.468	04:36
Air Blank		0.000	04:37
Control Test		0.076	04:37
Air Blank		0.000	04:38
Diagnostics Check		OK	04:38

Data Download Date/Time: 12/31/2019 19:20

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NICOLAS,MITA,
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: OLDHAM,TAYLOR
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 11/17/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: \
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:25

SUBJECT'S LAST NAME: LUSCOMBE

DATE/TIME OF TEST: 11/23/2019 05:55:54

Diagnostics Check		OK	05:46
Air Blank		0.000	05:46
Control Test		0.079	05:46
Air Blank		0.000	05:47
Subject Sample	# 1	0.113	05:47
Breath Volume		1.855	05:47
Air Blank		0.000	05:48
Air Blank		0.000	05:50
Subject Sample	# 2	0.136	05:51
Breath Volume		1.109	05:51
Air Blank		0.000	05:51
Air Blank		0.000	05:53
Subject Sample	# 3	0.126	05:54
Breath Volume		3.023	05:54
Air Blank		0.000	05:54
Control Test		0.079	05:55
Air Blank		0.000	05:55
Diagnostics Check		OK	05:55

Data Download Date/Time: 12/31/2019 19:20

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEYER,JEFFREY,D
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: DAWSON,M
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 11/17/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 10:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:02

SUBJECT'S LAST NAME: ARSENAULT

DATE/TIME OF TEST: 11/23/2019 11:31:42

Diagnostics Check		OK	11:24
Air Blank		0.000	11:25
Control Test		0.080	11:25
Air Blank		0.000	11:26
Subject Sample	# 1	0.101	11:27
Breath Volume		2.281	11:27
Air Blank		0.000	11:27
Air Blank		0.000	11:29
Subject Sample	# 2	0.100	11:30
Breath Volume		3.070	11:30
Air Blank		0.000	11:30
Control Test		0.079	11:31
Air Blank		0.000	11:31
Diagnostics Check		OK	11:31

Data Download Date/Time: 12/31/2019 19:20

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRIEMER,AMANDA,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: SPINA,T
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 11/17/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: PERKINS

DATE/TIME OF TEST: 11/25/2019 00:51:41

Diagnostics Check		OK	00:48	
Air Blank		0.000	00:49	
Control Test		0.080	00:49	
Air Blank		0.000	00:50	
Subject Sample	# 1		00:50	Subject Test Refused
Breath Volume		0.000	00:50	
Air Blank		0.000	00:50	
Control Test		0.080	00:51	
Air Blank		0.000	00:51	
Diagnostics Check		OK	00:51	

Data Download Date/Time: 12/31/2019 19:20

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIMA,MICHAEL,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: LIMA,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/17/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: WOOD

DATE/TIME OF TEST: 11/28/2019 02:02:02

Diagnostics Check		OK	01:55
Air Blank		0.000	01:56
Control Test		0.080	01:56
Air Blank		0.000	01:57
Subject Sample	# 1	0.135	01:57
Breath Volume		3.019	01:57
Air Blank		0.000	01:58
Air Blank		0.000	02:00
Subject Sample	# 2	0.138	02:00
Breath Volume		3.000	02:00
Air Blank		0.000	02:01
Control Test		0.079	02:01
Air Blank		0.000	02:01
Diagnostics Check		OK	02:02

Data Download Date/Time: 12/31/2019 19:20

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROUCH,J,L
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: ZAPATA,A
ARRESTING OFFICER AGENCY: Bunnell PD
LAST AGENCY INSPECTION: 11/17/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: KINARD

DATE/TIME OF TEST: 11/28/2019 19:56:45

Diagnostics Check		OK	19:50
Air Blank		0.000	19:51
Control Test		0.080	19:51
Air Blank		0.000	19:51
Subject Sample	# 1	0.197	19:52
Breath Volume		2.253	19:52
Air Blank		0.000	19:52
Air Blank		0.000	19:54
Subject Sample	# 2	0.193	19:55
Breath Volume		2.191	19:55
Air Blank		0.000	19:55
Control Test		0.079	19:56
Air Blank		0.000	19:56
Diagnostics Check		OK	19:56

Data Download Date/Time: 12/31/2019 19:20

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROUCH,J,L
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: DAWSON,M
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 11/17/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: ONEILL

DATE/TIME OF TEST: 11/28/2019 22:39:37

Diagnostics Check		OK	22:30
Air Blank		0.000	22:30
Control Test		0.080	22:31
Air Blank		0.000	22:31
Subject Sample	# 1	0.219	22:33
Breath Volume		1.371	22:33
Air Blank		0.000	22:34
Air Blank		0.000	22:35
Subject Sample	# 2	0.217	22:37
Breath Volume		1.769	22:37
Air Blank		0.000	22:38
Control Test		0.078	22:38
Air Blank		0.000	22:39
Diagnostics Check		OK	22:39

Data Download Date/Time: 12/31/2019 19:20

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROUCH,J,L
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: KOTOWSKI,P
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 11/17/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: SADA

DATE/TIME OF TEST: 11/29/2019 02:05:44

Diagnostics Check		OK	01:59	
Air Blank		0.000	02:00	
Control Test		0.079	02:00	
Air Blank		0.000	02:01	
Subject Sample	# 1		02:03	Subject Test Refused
Breath Volume		0.519	02:03	
Air Blank		0.000	02:04	
Control Test		0.079	02:05	
Air Blank		0.000	02:05	
Diagnostics Check		OK	02:05	

Data Download Date/Time: 12/31/2019 19:20

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONG,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: HILL,JOHN
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 11/17/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:41

SUBJECT'S LAST NAME: WEAVER

DATE/TIME OF TEST: 11/29/2019 19:08:25

Diagnostics Check		OK	19:05	
Air Blank		0.000	19:05	
Control Test		0.080	19:05	
Air Blank		0.000	19:06	
Subject Sample	# 1		19:07	Subject Test Refused
Breath Volume		0.000	19:07	
Air Blank		0.000	19:07	
Control Test		0.080	19:07	
Air Blank		0.000	19:08	
Diagnostics Check		OK	19:08	

Data Download Date/Time: 12/31/2019 19:20

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006621

INSTRUMENT REGISTERED TO: KENNEDY SPACE CENTER
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANKS,KENT,A
BREATH TEST OPERATOR AGENCY: Kennedy Space Center
ARRESTING OFFICER: JONES,NICHOLAS
ARRESTING OFFICER AGENCY: Kennedy Space Center
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: FANA

DATE/TIME OF TEST: 11/02/2019 04:32:54

Diagnostics Check		OK	04:26
Air Blank		0.000	04:26
Control Test		0.077	04:26
Air Blank		0.000	04:27
Subject Sample	# 1	0.150	04:28
Breath Volume		4.164	04:28
Air Blank		0.000	04:28
Air Blank		0.000	04:30
Subject Sample	# 2	0.149	04:31
Breath Volume		4.917	04:31
Air Blank		0.000	04:31
Control Test		0.076	04:32
Air Blank		0.000	04:32
Diagnostics Check		OK	04:32

Data Download Date/Time: 11/10/2019 15:43

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006621

INSTRUMENT REGISTERED TO: KENNEDY SPACE CENTER
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOYNER,KAREN,A
BREATH TEST OPERATOR AGENCY: Kennedy Space Center
ARRESTING OFFICER: LOPEZ,TIM
ARRESTING OFFICER AGENCY: OTHER Fed
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:12

SUBJECT'S LAST NAME: THOMS

DATE/TIME OF TEST: 11/10/2019 17:07:00

Diagnostics Check		OK	16:59
Air Blank		0.000	16:59
Control Test		0.078	17:00
Air Blank		0.000	17:00
Subject Sample	# 1	0.147	17:02
Breath Volume		2.128	17:02
Air Blank		0.000	17:03
Air Blank		0.000	17:04
Subject Sample	# 2	0.138	17:05
Breath Volume		1.582	17:05
Air Blank		0.000	17:06
Control Test		0.077	17:06
Air Blank		0.000	17:06
Diagnostics Check		OK	17:07

Data Download Date/Time: 11/11/2019 15:48

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006621

INSTRUMENT REGISTERED TO: KENNEDY SPACE CENTER
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCORMICK,PATRICK,K
BREATH TEST OPERATOR AGENCY: OTHER Fed
ARRESTING OFFICER: LOPEZ,TIM
ARRESTING OFFICER AGENCY: OTHER Fed
LAST AGENCY INSPECTION: 11/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:29

SUBJECT'S LAST NAME: SALERNO

DATE/TIME OF TEST: 11/23/2019 18:26:15

Diagnostics Check		OK	18:18
Air Blank		0.000	18:18
Control Test		0.078	18:19
Air Blank		0.000	18:19
Subject Sample	# 1	0.115	18:21
Breath Volume		4.500	18:21
Air Blank		0.000	18:22
Air Blank		0.000	18:23
Subject Sample	# 2	0.113	18:24
Breath Volume		3.371	18:24
Air Blank		0.000	18:25
Control Test		0.077	18:25
Air Blank		0.000	18:26
Diagnostics Check		OK	18:26

Data Download Date/Time: 12/12/2019 13:11

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,HALEY,M
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: FELIX,REYSON
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: SALAMON

DATE/TIME OF TEST: 11/01/2019 20:16:53

Diagnostics Check		OK	20:14	
Air Blank		0.000	20:14	
Control Test		0.078	20:14	
Air Blank		0.000	20:15	
Subject Sample	# 1		20:15	Subject Test Refused
Breath Volume		0.000	20:15	
Air Blank		0.000	20:15	
Control Test		0.079	20:16	
Air Blank		0.000	20:16	
Diagnostics Check		OK	20:16	

Data Download Date/Time: 11/26/2019 08:49

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARE,BRIAN,S
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: BISSOON,STEPHEN
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 06:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:50

SUBJECT'S LAST NAME: CHAMBLISS

DATE/TIME OF TEST: 11/09/2019 07:24:54

Diagnostics Check		OK	07:16
Air Blank		0.000	07:16
Control Test		0.080	07:16
Air Blank		0.000	07:17
Subject Sample	# 1	0.182	07:18
Breath Volume		1.976	07:18
Air Blank		0.000	07:18
Air Blank		0.000	07:20
Subject Sample	# 2	0.195	07:22
Breath Volume		2.335	07:22
Air Blank		0.000	07:23
Control Test		0.080	07:24
Air Blank		0.000	07:24
Diagnostics Check		OK	07:24

Data Download Date/Time: 11/26/2019 08:49

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNETTE,ASHLEY,N
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: HETTCHEN,JOHN
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: TRIEPER

DATE/TIME OF TEST: 11/14/2019 23:41:03

Diagnostics Check		OK	23:38	
Air Blank		0.000	23:38	
Control Test		0.078	23:38	
Air Blank		0.000	23:39	
Subject Sample	# 1		23:39	Subject Test Refused
Breath Volume		0.000	23:39	
Air Blank		0.000	23:40	
Control Test		0.080	23:40	
Air Blank		0.000	23:40	
Diagnostics Check		OK	23:41	

Data Download Date/Time: 11/26/2019 08:49

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZIADIE,ANDREW,C
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: HORNE,ASHTON
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 12:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:15

SUBJECT'S LAST NAME: COSTANZA

DATE/TIME OF TEST: 11/18/2019 13:56:27

Diagnostics Check		OK	13:53	
Air Blank		0.000	13:53	
Control Test		0.079	13:54	
Air Blank		0.000	13:54	
Subject Sample	# 1		13:55	Subject Test Refused
Breath Volume		0.000	13:55	
Air Blank		0.000	13:55	
Control Test		0.080	13:55	
Air Blank		0.000	13:56	
Diagnostics Check		OK	13:56	

Data Download Date/Time: 11/26/2019 08:49

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COON,REBECCA,L
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: BURNETTE,ASHLEY
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: ACKER

DATE/TIME OF TEST: 11/21/2019 03:19:48

Diagnostics Check		OK	03:16	
Air Blank		0.000	03:17	
Control Test		0.080	03:17	
Air Blank		0.000	03:17	
Subject Sample	# 1		03:18	Subject Test Refused
Breath Volume		0.000	03:18	
Air Blank		0.000	03:18	
Control Test		0.081	03:19	
Air Blank		0.000	03:19	
Diagnostics Check		OK	03:19	

Data Download Date/Time: 11/26/2019 08:49

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,HALEY,M
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: MCQUISTON,DEREK
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: DE VORE

DATE/TIME OF TEST: 11/23/2019 02:41:27

Diagnostics Check		OK	02:32
Air Blank		0.000	02:33
Control Test		0.079	02:33
Air Blank		0.000	02:33
Subject Sample	# 1	0.146	02:35
Breath Volume		1.824	02:35
Air Blank		0.000	02:36
Air Blank		0.000	02:38
Subject Sample	# 2	0.137	02:39
Breath Volume		1.699	02:39
Air Blank		0.000	02:40
Control Test		0.077	02:40
Air Blank		0.000	02:41
Diagnostics Check		OK	02:41

Data Download Date/Time: 11/26/2019 08:49

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNETTE,ASHLEY,N
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: CASAS,JAVIER
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: SALERNO

DATE/TIME OF TEST: 11/30/2019 01:54:11

Diagnostics Check		OK	01:46
Air Blank		0.000	01:47
Control Test		0.079	01:47
Air Blank		0.000	01:47
Subject Sample	# 1	0.301	01:48
Breath Volume		1.894	01:48
Air Blank		0.000	01:49
Air Blank		0.000	01:50
Subject Sample	# 2	0.306	01:52
Breath Volume		1.324	01:52
Air Blank		0.000	01:53
Control Test		0.079	01:53
Air Blank		0.000	01:54
Diagnostics Check		OK	01:54

Data Download Date/Time: 12/27/2019 09:03

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICCIARDI,AMANDA,L
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: VICK,BRYAN
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: WOLFLICK

DATE/TIME OF TEST: 11/30/2019 04:02:57

Diagnostics Check		OK	04:00	
Air Blank		0.000	04:00	
Control Test		0.079	04:00	
Air Blank		0.000	04:01	
Subject Sample	# 1		04:01	Subject Test Refused
Breath Volume		0.000	04:01	
Air Blank		0.000	04:02	
Control Test		0.079	04:02	
Air Blank		0.000	04:02	
Diagnostics Check		OK	04:02	

Data Download Date/Time: 12/27/2019 09:03

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006628

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,VICTOR,J
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: RIVERA,VICTOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 11/01/2019 04:13:36

Diagnostics Check		OK	04:06
Air Blank		0.000	04:07
Control Test		0.078	04:07
Air Blank		0.000	04:08
Subject Sample	# 1	0.131	04:08
Breath Volume		3.375	04:08
Air Blank		0.000	04:09
Air Blank		0.000	04:11
Subject Sample	# 2	0.128	04:12
Breath Volume		2.312	04:12
Air Blank		0.000	04:12
Control Test		0.078	04:12
Air Blank		0.000	04:13
Diagnostics Check		OK	04:13

Data Download Date/Time: 11/24/2019 22:34

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006628

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,VICTOR,J
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: RIVERA,VICTOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: JOUBRAN

DATE/TIME OF TEST: 11/03/2019 02:19:48

Diagnostics Check		OK	02:11
Air Blank		0.000	02:11
Control Test		0.079	02:11
Air Blank		0.000	02:12
Subject Sample	# 1	0.175	02:12
Breath Volume		1.628	02:12
Air Blank		0.000	02:13
Air Blank		0.000	02:15
Subject Sample	# 2	0.186	02:18
Breath Volume		1.617	02:18
Air Blank		0.000	02:18
Control Test		0.079	02:19
Air Blank		0.000	02:19
Diagnostics Check		OK	02:19

Data Download Date/Time: 11/24/2019 22:34

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006628

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARMAN,RICHARD,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HARMAN,RICHARD
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:11

SUBJECT'S LAST NAME: CONNOLLY

DATE/TIME OF TEST: 11/03/2019 05:42:58

Diagnostics Check		OK	05:35
Air Blank		0.000	05:35
Control Test		0.080	05:36
Air Blank		0.000	05:36
Subject Sample	# 1	0.159	05:37
Breath Volume		1.980	05:37
Air Blank		0.000	05:38
Air Blank		0.000	05:39
Subject Sample	# 2	0.165	05:41
Breath Volume		2.816	05:41
Air Blank		0.000	05:42
Control Test		0.080	05:42
Air Blank		0.000	05:42
Diagnostics Check		OK	05:42

Data Download Date/Time: 11/24/2019 22:34

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006628

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,VICTOR,J
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: RIVERA,VICTOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: KAYAD

DATE/TIME OF TEST: 11/08/2019 03:45:24

Diagnostics Check		OK	03:39
Air Blank		0.000	03:39
Control Test		0.078	03:39
Air Blank		0.000	03:40
Subject Sample	# 1	0.192	03:40
Breath Volume		1.976	03:40
Air Blank		0.000	03:41
Air Blank		0.000	03:43
Subject Sample	# 2	0.185	03:43
Breath Volume		2.195	03:43
Air Blank		0.000	03:44
Control Test		0.078	03:44
Air Blank		0.000	03:45
Diagnostics Check		OK	03:45

Data Download Date/Time: 11/24/2019 22:34

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006628

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,VICTOR,J
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: RIVERA,VICTOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: QUIEBRAS SANCHEZ

DATE/TIME OF TEST: 11/08/2019 23:42:34

Diagnostics Check		OK	23:36
Air Blank		0.000	23:36
Control Test		0.079	23:36
Air Blank		0.000	23:37
Subject Sample	# 1	0.183	23:37
Breath Volume		2.781	23:37
Air Blank		0.000	23:38
Air Blank		0.000	23:40
Subject Sample	# 2	0.178	23:41
Breath Volume		2.832	23:41
Air Blank		0.000	23:41
Control Test		0.078	23:41
Air Blank		0.000	23:42
Diagnostics Check		OK	23:42

Data Download Date/Time: 11/24/2019 22:34

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006628

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,VICTOR,J
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: RIVERA,VICTOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: ZERO TOL
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: CAMPOS

DATE/TIME OF TEST: 11/10/2019 03:04:38

Diagnostics Check		OK	02:58
Air Blank		0.000	02:58
Control Test		0.079	02:59
Air Blank		0.000	02:59
Subject Sample	# 1	0.050	03:00
Breath Volume		2.390	03:00
Air Blank		0.000	03:00
Air Blank		0.000	03:02
Subject Sample	# 2	0.049	03:03
Breath Volume		2.828	03:03
Air Blank		0.000	03:03
Control Test		0.079	03:03
Air Blank		0.000	03:04
Diagnostics Check		OK	03:04

Data Download Date/Time: 11/24/2019 22:34

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006628

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,VICTOR,J
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: RIVERA,VICTOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: CRUZ MALDONADO

DATE/TIME OF TEST: 11/17/2019 00:08:15

Diagnostics Check		OK	00:01
Air Blank		0.000	00:01
Control Test		0.080	00:02
Air Blank		0.000	00:02
Subject Sample	# 1	0.170	00:03
Breath Volume		2.011	00:03
Air Blank		0.000	00:04
Air Blank		0.000	00:05
Subject Sample	# 2	0.167	00:06
Breath Volume		1.816	00:06
Air Blank		0.000	00:07
Control Test		0.078	00:07
Air Blank		0.000	00:08
Diagnostics Check		OK	00:08

Data Download Date/Time: 11/24/2019 22:34

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006628

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,VICTOR,J
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: RIVERA,VICTOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 11/22/2019 04:03:58

Diagnostics Check		OK	03:57
Air Blank		0.000	03:58
Control Test		0.078	03:58
Air Blank		0.000	03:58
Subject Sample	# 1	0.180	03:59
Breath Volume		4.882	03:59
Air Blank		0.000	03:59
Air Blank		0.000	04:01
Subject Sample	# 2	0.186	04:02
Breath Volume		4.507	04:02
Air Blank		0.000	04:03
Control Test		0.078	04:03
Air Blank		0.000	04:03
Diagnostics Check		OK	04:03

Data Download Date/Time: 11/24/2019 22:34

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006628

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,VICTOR,J
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: RIVERA,VICTOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: AVILES TORRES

DATE/TIME OF TEST: 11/23/2019 01:29:49

Diagnostics Check		OK	01:22
Air Blank		0.000	01:23
Control Test		0.079	01:23
Air Blank		0.000	01:24
Subject Sample	# 1	0.192	01:24
Breath Volume		2.230	01:24
Air Blank		0.000	01:25
Air Blank		0.000	01:27
Subject Sample	# 2	0.184	01:28
Breath Volume		2.332	01:28
Air Blank		0.000	01:28
Control Test		0.079	01:29
Air Blank		0.000	01:29
Diagnostics Check		OK	01:29

Data Download Date/Time: 11/24/2019 22:34

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006628

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,VICTOR,J
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: RIVERA,VICTOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: PINEDA JR

DATE/TIME OF TEST: 11/28/2019 03:28:33

Diagnostics Check		OK	03:17	
Air Blank		0.000	03:17	
Control Test		0.078	03:17	
Air Blank		0.000	03:18	
Subject Sample	# 1		03:21	No Sample Provided
Breath Volume		0.000	03:21	
Air Blank		0.000	03:22	
Air Blank		0.000	03:24	
Subject Sample	# 2		03:27	No Sample Provided
Breath Volume		0.000	03:27	
Air Blank		0.000	03:27	
Control Test		0.077	03:27	
Air Blank		0.000	03:28	
Diagnostics Check		OK	03:28	

Data Download Date/Time: 12/11/2019 23:13

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006628

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,VICTOR,J
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: RIVERA,VICTOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: PALOMAR

DATE/TIME OF TEST: 11/30/2019 04:48:35

Diagnostics Check		OK	04:41
Air Blank		0.000	04:42
Control Test		0.079	04:42
Air Blank		0.000	04:43
Subject Sample	# 1	0.161	04:43
Breath Volume		3.789	04:43
Air Blank		0.000	04:44
Air Blank		0.000	04:46
Subject Sample	# 2	0.158	04:46
Breath Volume		4.015	04:46
Air Blank		0.000	04:47
Control Test		0.078	04:47
Air Blank		0.000	04:48
Diagnostics Check		OK	04:48

Data Download Date/Time: 12/11/2019 23:13

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006629

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAWKINS,DEBORAH,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HAWKINS,DEBORAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1791908A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 11/02/2019 18:11:59

Diagnostics Check	OK	18:11	
Air Blank		18:11	RFI Detect
Air Blank	0.000	18:11	

Data Download Date/Time: 11/24/2019 22:42

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006629

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAWKINS,DEBORAH,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HAWKINS,DEBORAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1791908A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 11/02/2019 18:20:08

Diagnostics Check		OK	18:13
Air Blank		0.000	18:14
Control Test		0.077	18:14
Air Blank		0.000	18:15
Subject Sample	# 1	0.215	18:15
Breath Volume		3.800	18:15
Air Blank		0.000	18:16
Air Blank		0.000	18:18
Subject Sample	# 2	0.207	18:18
Breath Volume		4.445	18:18
Air Blank		0.000	18:19
Control Test		0.078	18:19
Air Blank		0.000	18:20
Diagnostics Check		OK	18:20

Data Download Date/Time: 11/24/2019 22:42

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: GOLDSBY

DATE/TIME OF TEST: 11/01/2019 02:29:53

Diagnostics Check		OK	02:23
Air Blank		0.000	02:23
Control Test		0.080	02:24
Air Blank		0.000	02:24
Subject Sample	# 1	0.047	02:25
Breath Volume		4.550	02:25
Air Blank		0.000	02:26
Air Blank		0.000	02:27
Subject Sample	# 2	0.039	02:28
Breath Volume		1.425	02:28
Air Blank		0.000	02:28
Control Test		0.079	02:29
Air Blank		0.000	02:29
Diagnostics Check		OK	02:29

Data Download Date/Time: 11/18/2019 13:08

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: POUGH

DATE/TIME OF TEST: 11/02/2019 01:54:56

Diagnostics Check		OK	01:48
Air Blank		0.000	01:49
Control Test		0.080	01:49
Air Blank		0.000	01:50
Subject Sample	# 1	0.000	01:50
Breath Volume		1.222	01:50
Air Blank		0.000	01:51
Air Blank		0.000	01:53
Subject Sample	# 2	0.000	01:53
Breath Volume		1.679	01:53
Air Blank		0.000	01:54
Control Test		0.080	01:54
Air Blank		0.000	01:54
Diagnostics Check		OK	01:54

Data Download Date/Time: 11/18/2019 13:08

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGBY,CHARLES,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BIGBY,CHARLES
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: VASQUEZ IGLESIAS

DATE/TIME OF TEST: 11/02/2019 22:42:03

Diagnostics Check		OK	22:38	
Air Blank		0.000	22:39	
Control Test		0.080	22:39	
Air Blank		0.000	22:40	
Subject Sample	# 1		22:40	Subject Test Refused
Breath Volume		0.000	22:40	
Air Blank		0.000	22:41	
Control Test		0.080	22:41	
Air Blank		0.000	22:41	
Diagnostics Check		OK	22:42	

Data Download Date/Time: 11/18/2019 13:08

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: BERRUETA

DATE/TIME OF TEST: 11/04/2019 00:50:51

Diagnostics Check		OK	00:41	
Air Blank		0.000	00:41	
Control Test		0.080	00:41	
Air Blank		0.000	00:42	
Subject Sample	# 1		00:45	No Sample Provided
Breath Volume		0.000	00:45	
Air Blank		0.000	00:46	
Air Blank		0.000	00:48	
Subject Sample	# 2		00:49	Subject Test Refused
Breath Volume		0.000	00:49	
Air Blank		0.000	00:49	
Control Test		0.080	00:50	
Air Blank		0.000	00:50	
Diagnostics Check		OK	00:50	

Data Download Date/Time: 11/18/2019 13:08

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: HUERTA

DATE/TIME OF TEST: 11/07/2019 02:07:40

Diagnostics Check		OK	02:01
Air Blank		0.000	02:01
Control Test		0.079	02:02
Air Blank		0.000	02:02
Subject Sample	# 1	0.000	02:03
Breath Volume		1.914	02:03
Air Blank		0.000	02:03
Air Blank		0.000	02:05
Subject Sample	# 2	0.000	02:06
Breath Volume		2.628	02:06
Air Blank		0.000	02:06
Control Test		0.080	02:07
Air Blank		0.000	02:07
Diagnostics Check		OK	02:07

Data Download Date/Time: 11/18/2019 13:08

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGBY,CHARLES,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BIGBY,CHARLES
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 18:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:07

SUBJECT'S LAST NAME: ASHBY JR

DATE/TIME OF TEST: 11/07/2019 19:35:39

Diagnostics Check		OK	19:32	
Air Blank		0.000	19:32	
Control Test		0.080	19:33	
Air Blank		0.000	19:33	
Subject Sample	# 1		19:33	RFI Detect
Breath Volume		1.683	19:33	
Air Blank		0.000	19:34	
Control Test		0.080	19:35	
Air Blank		0.000	19:35	
Diagnostics Check		OK	19:35	

Data Download Date/Time: 11/18/2019 13:08

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGBY,CHARLES,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BIGBY,CHARLES
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 18:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:07

SUBJECT'S LAST NAME: ASHBY JR

DATE/TIME OF TEST: 11/07/2019 19:44:11

Diagnostics Check		OK	19:37
Air Blank		0.000	19:38
Control Test		0.079	19:38
Air Blank		0.000	19:39
Subject Sample	# 1	0.176	19:39
Breath Volume		1.371	19:39
Air Blank		0.000	19:40
Air Blank		0.000	19:42
Subject Sample	# 2	0.172	19:42
Breath Volume		1.746	19:42
Air Blank		0.000	19:43
Control Test		0.079	19:43
Air Blank		0.000	19:44
Diagnostics Check		OK	19:44

Data Download Date/Time: 11/18/2019 13:08

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BIGBY,CHARLES
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: KOS

DATE/TIME OF TEST: 11/08/2019 01:59:45

Diagnostics Check		OK	01:53
Air Blank		0.000	01:53
Control Test		0.079	01:54
Air Blank		0.000	01:54
Subject Sample	# 1	0.167	01:55
Breath Volume		2.906	01:55
Air Blank		0.000	01:55
Air Blank		0.000	01:57
Subject Sample	# 2	0.169	01:58
Breath Volume		2.875	01:58
Air Blank		0.000	01:58
Control Test		0.079	01:59
Air Blank		0.000	01:59
Diagnostics Check		OK	01:59

Data Download Date/Time: 11/18/2019 13:08

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: MILLS

DATE/TIME OF TEST: 11/09/2019 23:44:22

Diagnostics Check		OK	23:34
Air Blank		0.000	23:34
Control Test		0.081	23:35
Air Blank		0.000	23:35
Subject Sample	# 1	0.000	23:38
Breath Volume		1.156	23:38
Air Blank		0.000	23:39
Air Blank		0.000	23:41
Subject Sample	# 2	0.000	23:42
Breath Volume		1.265	23:42
Air Blank		0.000	23:43
Control Test		0.080	23:43
Air Blank		0.000	23:44
Diagnostics Check		OK	23:44

Data Download Date/Time: 11/18/2019 13:08

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,CODY,R
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BARRETT,CODY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: STEVELEY

DATE/TIME OF TEST: 11/10/2019 23:11:43

Diagnostics Check		OK	23:00	
Air Blank		0.000	23:01	
Control Test		0.080	23:01	
Air Blank		0.000	23:01	
Subject Sample	# 1	0.155	23:05	Volume Not Met
Breath Volume		1.046	23:05	
Air Blank		0.000	23:05	
Air Blank		0.000	23:07	
Subject Sample	# 2	0.153	23:10	Volume Not Met
Breath Volume		0.886	23:10	
Air Blank			23:11	Sequence Aborted
Air Blank		0.000	23:11	

Data Download Date/Time: 11/18/2019 13:08

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,CODY,R
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BARRETT,CODY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: STEVELEY

DATE/TIME OF TEST: 11/10/2019 23:26:45

Diagnostics Check		OK	23:15	
Air Blank		0.000	23:16	
Control Test		0.079	23:16	
Air Blank		0.000	23:16	
Subject Sample	# 1	0.140	23:20	Volume Not Met
Breath Volume		0.015	23:20	
Air Blank		0.000	23:20	
Air Blank		0.000	23:22	
Subject Sample	# 2		23:25	Subject Test Refused
Breath Volume		0.824	23:25	
Air Blank		0.000	23:25	
Control Test		0.077	23:26	
Air Blank		0.000	23:26	
Diagnostics Check		OK	23:26	

Data Download Date/Time: 11/18/2019 13:08

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 11/12/2019 00:23:16

Diagnostics Check		OK	00:16
Air Blank		0.000	00:17
Control Test		0.079	00:17
Air Blank		0.000	00:17
Subject Sample	# 1	0.082	00:18
Breath Volume		4.578	00:18
Air Blank		0.000	00:19
Air Blank		0.000	00:21
Subject Sample	# 2	0.076	00:21
Breath Volume		2.582	00:21
Air Blank		0.000	00:22
Control Test		0.080	00:22
Air Blank		0.000	00:23
Diagnostics Check		OK	00:23

Data Download Date/Time: 11/18/2019 13:08

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGBY,CHARLES,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BIGBY,CHARLES
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 11/12/2019 01:52:34

Diagnostics Check		OK	01:46
Air Blank		0.000	01:47
Control Test		0.080	01:47
Air Blank		0.000	01:47
Subject Sample	# 1	0.000	01:48
Breath Volume		1.464	01:48
Air Blank		0.000	01:48
Air Blank		0.000	01:50
Subject Sample	# 2	0.000	01:51
Breath Volume		2.589	01:51
Air Blank		0.000	01:51
Control Test		0.079	01:51
Air Blank		0.000	01:52
Diagnostics Check		OK	01:52

Data Download Date/Time: 11/18/2019 13:08

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KING,GRACE,N
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: CUMMINGS,GARY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 04:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:12

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 11/12/2019 05:36:36

Diagnostics Check		OK	05:33	
Air Blank		0.000	05:34	
Control Test		0.079	05:34	
Air Blank		0.000	05:34	
Subject Sample	# 1		05:35	Subject Test Refused
Breath Volume		0.000	05:35	
Air Blank		0.000	05:35	
Control Test		0.080	05:36	
Air Blank		0.000	05:36	
Diagnostics Check		OK	05:36	

Data Download Date/Time: 11/18/2019 13:08

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KING,GRACE,N
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: CUMMINGS,GARY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 04:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:12

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 11/12/2019 05:57:09

Diagnostics Check		OK	05:50
Air Blank		0.000	05:50
Control Test		0.080	05:50
Air Blank		0.000	05:51
Subject Sample	# 1	0.110	05:52
Breath Volume		1.621	05:52
Air Blank		0.000	05:53
Air Blank		0.000	05:54
Subject Sample	# 2	0.107	05:55
Breath Volume		1.582	05:55
Air Blank		0.000	05:56
Control Test		0.077	05:56
Air Blank		0.000	05:57
Diagnostics Check		OK	05:57

Data Download Date/Time: 11/18/2019 13:08

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGBY,CHARLES,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BIGBY,CHARLES
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:33

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 11/12/2019 22:09:37

Diagnostics Check		OK	21:53	
Air Blank		0.000	21:53	
Control Test		0.079	21:54	
Air Blank		0.000	21:54	
Subject Sample	# 1	0.000	21:57	Volume Not Met
Breath Volume		0.250	21:57	
Air Blank		0.000	21:58	
Air Blank		0.000	22:00	
Subject Sample	# 2	0.008	22:02	
Breath Volume		1.933	22:02	
Air Blank		0.000	22:03	
Air Blank		0.000	22:05	
Subject Sample	# 3	0.007	22:08	
Breath Volume		1.476	22:08	
Air Blank		0.000	22:08	
Control Test		0.079	22:09	
Air Blank		0.000	22:09	
Diagnostics Check		OK	22:09	

Data Download Date/Time: 11/18/2019 13:08

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 18:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:41

SUBJECT'S LAST NAME: ELDERT

DATE/TIME OF TEST: 11/13/2019 19:07:47

Diagnostics Check		OK	19:01
Air Blank		0.000	19:02
Control Test		0.080	19:02
Air Blank		0.000	19:03
Subject Sample	# 1	0.000	19:03
Breath Volume		1.429	19:03
Air Blank		0.000	19:04
Air Blank		0.000	19:06
Subject Sample	# 2	0.000	19:06
Breath Volume		1.453	19:06
Air Blank		0.000	19:06
Control Test		0.077	19:07
Air Blank		0.000	19:07
Diagnostics Check		OK	19:07

Data Download Date/Time: 11/18/2019 13:08

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 18:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:39

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 11/15/2019 19:05:46

Diagnostics Check		OK	18:59
Air Blank		0.000	18:59
Control Test		0.081	19:00
Air Blank		0.000	19:00
Subject Sample	# 1	0.133	19:01
Breath Volume		3.894	19:01
Air Blank		0.000	19:01
Air Blank		0.000	19:03
Subject Sample	# 2	0.113	19:04
Breath Volume		1.488	19:04
Air Blank		0.000	19:04
Control Test		0.079	19:05
Air Blank		0.000	19:05
Diagnostics Check		OK	19:05

Data Download Date/Time: 11/18/2019 13:08

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: RESENDIZ

DATE/TIME OF TEST: 11/15/2019 22:55:36

Diagnostics Check		OK	22:49
Air Blank		0.000	22:49
Control Test		0.079	22:50
Air Blank		0.000	22:50
Subject Sample	# 1	0.221	22:51
Breath Volume		1.769	22:51
Air Blank		0.000	22:51
Air Blank		0.000	22:53
Subject Sample	# 2	0.226	22:53
Breath Volume		1.882	22:53
Air Blank		0.000	22:54
Control Test		0.079	22:54
Air Blank		0.000	22:55
Diagnostics Check		OK	22:55

Data Download Date/Time: 11/18/2019 13:08

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: GRAY

DATE/TIME OF TEST: 11/16/2019 02:59:23

Diagnostics Check		OK	02:52
Air Blank		0.000	02:53
Control Test		0.080	02:53
Air Blank		0.000	02:54
Subject Sample	# 1	0.193	02:54
Breath Volume		3.035	02:54
Air Blank		0.000	02:55
Air Blank		0.000	02:57
Subject Sample	# 2	0.196	02:57
Breath Volume		2.343	02:57
Air Blank		0.000	02:58
Control Test		0.080	02:58
Air Blank		0.000	02:59
Diagnostics Check		OK	02:59

Data Download Date/Time: 11/18/2019 13:08

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: DIAZ ROMERO

DATE/TIME OF TEST: 11/18/2019 02:25:44

Diagnostics Check		OK	02:19
Air Blank		0.000	02:19
Control Test		0.080	02:20
Air Blank		0.000	02:20
Subject Sample	# 1	0.175	02:21
Breath Volume		2.039	02:21
Air Blank		0.000	02:21
Air Blank		0.000	02:23
Subject Sample	# 2	0.178	02:24
Breath Volume		1.359	02:24
Air Blank		0.000	02:24
Control Test		0.080	02:25
Air Blank		0.000	02:25
Diagnostics Check		OK	02:25

Data Download Date/Time: 11/18/2019 13:08

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: VASQUEZ MENDOZA

DATE/TIME OF TEST: 11/21/2019 23:20:33

Diagnostics Check		OK	23:14
Air Blank		0.000	23:14
Control Test		0.080	23:14
Air Blank		0.000	23:15
Subject Sample	# 1	0.109	23:15
Breath Volume		1.390	23:15
Air Blank		0.000	23:16
Air Blank		0.000	23:18
Subject Sample	# 2	0.102	23:18
Breath Volume		1.566	23:18
Air Blank		0.000	23:19
Control Test		0.079	23:19
Air Blank		0.000	23:20
Diagnostics Check		OK	23:20

Data Download Date/Time: 12/17/2019 07:46

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 11/21/2019 23:48:14

Diagnostics Check		OK	23:43	
Air Blank		0.000	23:44	
Control Test		0.080	23:44	
Air Blank		0.000	23:45	
Subject Sample	# 1		23:46	Subject Test Refused
Breath Volume		0.000	23:46	
Air Blank		0.000	23:47	
Control Test		0.080	23:47	
Air Blank		0.000	23:48	
Diagnostics Check		OK	23:48	

Data Download Date/Time: 12/17/2019 07:46

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 11/22/2019 00:01:38

Diagnostics Check		OK	23:49	
Air Blank		0.000	23:50	
Control Test		0.080	23:50	
Air Blank		0.000	23:51	
Subject Sample	# 2	0.150	23:54	Volume Not Met
Breath Volume		1.058	23:54	
Air Blank		0.000	23:55	
Air Blank		0.000	23:56	
Subject Sample	# 1	0.150	00:00	Volume Not Met
Breath Volume		0.015	00:00	
Air Blank		0.000	00:00	
Control Test		0.078	00:01	
Air Blank		0.000	00:01	
Diagnostics Check		OK	00:01	

Data Download Date/Time: 12/17/2019 07:46

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 11/22/2019 00:12:59

Diagnostics Check		OK	00:05
Air Blank		0.000	00:05
Control Test		0.079	00:06
Air Blank		0.000	00:06
Subject Sample	# 1	0.162	00:07
Breath Volume		2.253	00:07
Air Blank		0.000	00:07
Air Blank		0.000	00:09
Subject Sample	# 2	0.157	00:11
Breath Volume		1.980	00:11
Air Blank		0.000	00:12
Control Test		0.080	00:12
Air Blank		0.000	00:12
Diagnostics Check		OK	00:12

Data Download Date/Time: 12/17/2019 07:46

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,CODY,R
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BARRETT,CODY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: WEATHERLY

DATE/TIME OF TEST: 11/25/2019 02:51:23

Diagnostics Check		OK	02:48	
Air Blank		0.000	02:48	
Control Test		0.081	02:48	
Air Blank		0.000	02:49	
Subject Sample	# 1		02:49	RFI Detect
Breath Volume		0.000	02:49	
Air Blank		0.000	02:50	
Control Test		0.081	02:50	
Air Blank			02:50	RFI Detect
Air Blank		0.000	02:51	

Data Download Date/Time: 12/17/2019 07:46

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,CODY,R
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BARRETT,CODY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: WEATHERLY

DATE/TIME OF TEST: 11/25/2019 02:59:44

Diagnostics Check		OK	02:53
Air Blank		0.000	02:53
Control Test		0.080	02:54
Air Blank		0.000	02:54
Subject Sample	# 1	0.000	02:55
Breath Volume		2.269	02:55
Air Blank		0.000	02:55
Air Blank		0.000	02:57
Subject Sample	# 2	0.000	02:58
Breath Volume		1.589	02:58
Air Blank		0.000	02:58
Control Test		0.078	02:59
Air Blank		0.000	02:59
Diagnostics Check		OK	02:59

Data Download Date/Time: 12/17/2019 07:46

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGBY,CHARLES,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BIGBY,CHARLES
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: TRACEY

DATE/TIME OF TEST: 11/26/2019 00:09:25

Diagnostics Check		OK	00:02
Air Blank		0.000	00:03
Control Test		0.079	00:03
Air Blank		0.000	00:04
Subject Sample	# 1	0.131	00:04
Breath Volume		2.058	00:04
Air Blank		0.000	00:05
Air Blank		0.000	00:07
Subject Sample	# 2	0.123	00:07
Breath Volume		1.867	00:07
Air Blank		0.000	00:08
Control Test		0.079	00:08
Air Blank		0.000	00:09
Diagnostics Check		OK	00:09

Data Download Date/Time: 12/17/2019 07:47

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: CASTILLANOS REYES

DATE/TIME OF TEST: 11/28/2019 01:24:03

Diagnostics Check		OK	01:17
Air Blank		0.000	01:18
Control Test		0.081	01:18
Air Blank		0.000	01:19
Subject Sample	# 1	0.105	01:19
Breath Volume		2.535	01:19
Air Blank		0.000	01:20
Air Blank		0.000	01:22
Subject Sample	# 2	0.106	01:22
Breath Volume		2.769	01:22
Air Blank		0.000	01:23
Control Test		0.079	01:23
Air Blank		0.000	01:23
Diagnostics Check		OK	01:24

Data Download Date/Time: 12/17/2019 07:47

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:53

SUBJECT'S LAST NAME: ANDINO OCAMPO

DATE/TIME OF TEST: 11/28/2019 21:19:45

Diagnostics Check		OK	21:13
Air Blank		0.000	21:13
Control Test		0.081	21:13
Air Blank		0.000	21:14
Subject Sample	# 1	0.177	21:15
Breath Volume		2.664	21:15
Air Blank		0.000	21:15
Air Blank		0.000	21:17
Subject Sample	# 2	0.178	21:18
Breath Volume		2.746	21:18
Air Blank		0.000	21:18
Control Test		0.080	21:19
Air Blank		0.000	21:19
Diagnostics Check		OK	21:19

Data Download Date/Time: 12/17/2019 07:47

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: LARREYNAGA GONZALEZ

DATE/TIME OF TEST: 11/30/2019 23:59:03

Diagnostics Check		OK	23:52
Air Blank		0.000	23:53
Control Test		0.080	23:53
Air Blank		0.000	23:54
Subject Sample	# 1	0.071	23:54
Breath Volume		1.386	23:54
Air Blank		0.000	23:55
Air Blank		0.000	23:56
Subject Sample	# 2	0.072	23:57
Breath Volume		2.136	23:57
Air Blank		0.000	23:58
Control Test		0.080	23:58
Air Blank		0.000	23:58
Diagnostics Check		OK	23:59

Data Download Date/Time: 12/17/2019 07:47

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GWODZ,OFC,R
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: MOORE,ADAM
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 11/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111479
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: HUGGINS

DATE/TIME OF TEST: 11/19/2019 00:12:40

Diagnostics Check		OK	00:09	
Air Blank		0.000	00:10	
Control Test		0.079	00:10	
Air Blank		0.000	00:11	
Subject Sample	# 1		00:11	Subject Test Refused
Breath Volume		0.000	00:11	
Air Blank		0.000	00:11	
Control Test		0.079	00:12	
Air Blank		0.000	00:12	
Diagnostics Check		OK	00:12	

Data Download Date/Time: 12/16/2019 11:43

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGBY,CHARLES,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BIGBY,CHARLES
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 962587
DRY GAS EXPIRY DATE: 06/04/2021
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: AGUIRRE-LOPEZ

DATE/TIME OF TEST: 11/22/2019 02:01:48

Diagnostics Check		OK	01:57	
Air Blank		0.000	01:58	
Control Test		0.079	01:58	
Air Blank		0.000	01:59	
Subject Sample	# 1		02:00	Subject Test Refused
Breath Volume		0.000	02:00	
Air Blank		0.000	02:00	
Control Test		0.080	02:01	
Air Blank		0.000	02:01	
Diagnostics Check		OK	02:01	

Data Download Date/Time: 12/16/2019 11:43

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORD,JOSHUA,N
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LORD,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 11/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 962587
DRY GAS EXPIRY DATE: 06/04/2021
VIOLATION CODE: DUI
ARREST TIME: 21:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: POMPILUS

DATE/TIME OF TEST: 11/23/2019 22:32:26

Diagnostics Check		OK	22:20	
Air Blank		0.000	22:21	
Control Test		0.079	22:21	
Air Blank		0.000	22:22	
Subject Sample	# 1	0.100	22:25	Volume Not Met
Breath Volume		0.058	22:25	
Air Blank		0.000	22:25	
Air Blank		0.000	22:27	
Subject Sample	# 2		22:30	No Sample Provided
Breath Volume		0.000	22:30	
Air Blank		0.000	22:31	
Control Test		0.078	22:31	
Air Blank		0.000	22:32	
Diagnostics Check		OK	22:32	

Data Download Date/Time: 12/16/2019 11:43

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GWODZ,OFC,R
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: MOORE,ADAM
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 11/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: CULHANE

DATE/TIME OF TEST: 11/23/2019 23:32:06

Diagnostics Check		OK	23:28	
Air Blank		0.000	23:29	
Control Test		0.079	23:29	
Air Blank		0.000	23:29	
Subject Sample	# 1		23:30	Subject Test Refused
Breath Volume		0.000	23:30	
Air Blank		0.000	23:31	
Control Test		0.080	23:31	
Air Blank		0.000	23:32	
Diagnostics Check		OK	23:32	

Data Download Date/Time: 12/16/2019 11:43

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GWODZ,OFC,R
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: WENTLAND,ALEX
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 11/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111527
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: SCHEDENECK

DATE/TIME OF TEST: 11/24/2019 01:54:58

Diagnostics Check	OK	01:53	
Air Blank	0.000	01:53	
Control Test	0.079	01:54	
Air Blank		01:54	Improper Sample
Air Blank	0.000	01:54	

Data Download Date/Time: 12/16/2019 11:43

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GWODZ,OFC,R
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: WENTLAND,ALEX
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 11/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111527
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: SCHEDENECK

DATE/TIME OF TEST: 11/24/2019 02:04:22

Diagnostics Check		OK	02:01	
Air Blank		0.000	02:01	
Control Test		0.079	02:01	
Air Blank		0.000	02:02	
Subject Sample	# 1		02:03	Subject Test Refused
Breath Volume		0.000	02:03	
Air Blank		0.000	02:03	
Control Test		0.079	02:03	
Air Blank		0.000	02:04	
Diagnostics Check		OK	02:04	

Data Download Date/Time: 12/16/2019 11:43

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACEVEDO VALDES,CRISTIAN,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: MOORE,JASON
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 11/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 962587
DRY GAS EXPIRY DATE: 06/04/2021
VIOLATION CODE: DUI
ARREST TIME: 12:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: MOLINA

DATE/TIME OF TEST: 11/24/2019 04:30:28

Diagnostics Check		OK	04:27	
Air Blank		0.000	04:27	
Control Test		0.079	04:28	
Air Blank		0.000	04:28	
Subject Sample	# 1		04:29	Subject Test Refused
Breath Volume		0.000	04:29	
Air Blank		0.000	04:29	
Control Test		0.079	04:29	
Air Blank		0.000	04:30	
Diagnostics Check		OK	04:30	

Data Download Date/Time: 12/16/2019 11:43

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAUGHN,RYAN,
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: VAUGHN,RYAN
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 11/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 962587
DRY GAS EXPIRY DATE: 06/04/2021
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: DRAKE

DATE/TIME OF TEST: 11/26/2019 23:41:07

Diagnostics Check		OK	23:38	
Air Blank		0.000	23:38	
Control Test		0.080	23:39	
Air Blank		0.000	23:39	
Subject Sample	# 1		23:39	Subject Test Refused
Breath Volume		0.000	23:39	
Air Blank		0.000	23:40	
Control Test		0.080	23:40	
Air Blank		0.000	23:41	
Diagnostics Check		OK	23:41	

Data Download Date/Time: 12/16/2019 11:43

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAUGHN,RYAN,
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: MOORE,ADAM
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 11/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 962587
DRY GAS EXPIRY DATE: 06/04/2021
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: BOYD

DATE/TIME OF TEST: 11/28/2019 03:32:11

Diagnostics Check		OK	03:25
Air Blank		0.000	03:26
Control Test		0.080	03:26
Air Blank		0.000	03:26
Subject Sample	# 1	0.223	03:27
Breath Volume		4.601	03:27
Air Blank		0.000	03:28
Air Blank		0.000	03:29
Subject Sample	# 2	0.218	03:30
Breath Volume		4.425	03:30
Air Blank		0.000	03:31
Control Test		0.079	03:31
Air Blank		0.000	03:32
Diagnostics Check		OK	03:32

Data Download Date/Time: 12/16/2019 11:43

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOMBERLIN,KEVIN,W
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: TOMBERLIN,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 11/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 962587
DRY GAS EXPIRY DATE: 06/04/2021
VIOLATION CODE: DUI
ARREST TIME: 06:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: JOHNS

DATE/TIME OF TEST: 11/28/2019 09:40:39

Diagnostics Check		OK	09:37	
Air Blank		0.000	09:38	
Control Test		0.079	09:38	
Air Blank		0.000	09:39	
Subject Sample	# 1		09:39	Subject Test Refused
Breath Volume		0.000	09:39	
Air Blank		0.000	09:39	
Control Test		0.079	09:40	
Air Blank		0.000	09:40	
Diagnostics Check		OK	09:40	

Data Download Date/Time: 12/16/2019 11:43

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORD,JOSHUA,N
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LORD,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 11/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 962587
DRY GAS EXPIRY DATE: 06/04/2021
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: CRUM

DATE/TIME OF TEST: 11/28/2019 23:08:21

Diagnostics Check		OK	23:00
Air Blank		0.000	23:01
Control Test		0.079	23:01
Air Blank		0.000	23:02
Subject Sample	# 1	0.164	23:02
Breath Volume		1.609	23:02
Air Blank		0.000	23:03
Air Blank		0.000	23:05
Subject Sample	# 2	0.166	23:06
Breath Volume		1.246	23:06
Air Blank		0.000	23:07
Control Test		0.079	23:07
Air Blank		0.000	23:08
Diagnostics Check		OK	23:08

Data Download Date/Time: 12/16/2019 11:43

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GWODZ,OFC,R
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: MOORE,ADAM
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 11/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: TABAR

DATE/TIME OF TEST: 11/29/2019 00:55:08

Diagnostics Check		OK	00:52	
Air Blank		0.000	00:52	
Control Test		0.079	00:52	
Air Blank		0.000	00:53	
Subject Sample	# 1		00:53	Subject Test Refused
Breath Volume		0.000	00:53	
Air Blank		0.000	00:54	
Control Test		0.079	00:54	
Air Blank		0.000	00:55	
Diagnostics Check		OK	00:55	

Data Download Date/Time: 12/16/2019 11:43

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GWODZ,OFC,R
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: MOORE,ADAM
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 11/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 11/29/2019 04:41:38

Diagnostics Check		OK	04:28	
Air Blank		0.000	04:29	
Control Test		0.079	04:29	
Air Blank		0.000	04:30	
Subject Sample	# 1	0.129	04:33	Volume Not Met
Breath Volume		0.589	04:33	
Air Blank		0.000	04:33	
Air Blank		0.000	04:35	
Subject Sample	# 2	0.128	04:36	
Breath Volume		1.152	04:36	
Air Blank		0.000	04:37	
Air Blank		0.000	04:38	
Subject Sample	# 3	0.121	04:39	
Breath Volume		1.515	04:39	
Air Blank		0.000	04:40	
Control Test		0.077	04:41	
Air Blank		0.000	04:41	
Diagnostics Check		OK	04:41	

Data Download Date/Time: 12/16/2019 11:43

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006637

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALFY,RACHEL,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: PALFY,RACHEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: RODRIGUEZ MANOLO

DATE/TIME OF TEST: 11/03/2019 02:47:34

Diagnostics Check		OK	02:40
Air Blank		0.000	02:40
Control Test		0.079	02:41
Air Blank		0.000	02:41
Subject Sample	# 1	0.123	02:42
Breath Volume		1.898	02:42
Air Blank		0.000	02:43
Air Blank		0.000	02:44
Subject Sample	# 2	0.137	02:45
Breath Volume		1.812	02:45
Air Blank		0.000	02:46
Control Test		0.079	02:46
Air Blank		0.000	02:47
Diagnostics Check		OK	02:47

Data Download Date/Time: 11/28/2019 09:51

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

INSTRUMENT REGISTERED TO: HOLMES BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HURT,ALEXANDER,C
BREATH TEST OPERATOR AGENCY: Holmes Beach PD
ARRESTING OFFICER: HURT,ALEXANDER
ARRESTING OFFICER AGENCY: Holmes Beach PD
LAST AGENCY INSPECTION: 10/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175970
DRY GAS EXPIRY DATE: 08/02/2022
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: PEACHER

DATE/TIME OF TEST: 11/02/2019 03:25:36

Diagnostics Check		OK	03:13
Air Blank		0.000	03:14
Control Test		0.080	03:14
Air Blank		0.000	03:15
Subject Sample	# 1	0.194	03:16
Breath Volume		1.609	03:16
Air Blank		0.000	03:16
Air Blank		0.000	03:18
Subject Sample	# 2	0.155	03:20
Breath Volume		1.152	03:20
Air Blank		0.000	03:21
Air Blank		0.000	03:23
Subject Sample	# 3	0.178	03:23
Breath Volume		1.812	03:23
Air Blank		0.000	03:24
Control Test		0.076	03:25
Air Blank		0.000	03:25
Diagnostics Check		OK	03:25

Data Download Date/Time: 11/18/2019 10:01

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

INSTRUMENT REGISTERED TO: HOLMES BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HURT,ALEXANDER,C
BREATH TEST OPERATOR AGENCY: Holmes Beach PD
ARRESTING OFFICER: HURT,ALEXANDER
ARRESTING OFFICER AGENCY: Holmes Beach PD
LAST AGENCY INSPECTION: 10/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175970
DRY GAS EXPIRY DATE: 08/02/2022
VIOLATION CODE: DUI
ARREST TIME: 18:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:51

SUBJECT'S LAST NAME: EHRLER-MANZUR

DATE/TIME OF TEST: 11/07/2019 20:19:02

Diagnostics Check		OK	20:13
Air Blank		0.000	20:13
Control Test		0.080	20:13
Air Blank		0.000	20:14
Subject Sample	# 1	0.000	20:14
Breath Volume		3.425	20:14
Air Blank		0.000	20:15
Air Blank		0.000	20:17
Subject Sample	# 2	0.000	20:17
Breath Volume		3.343	20:17
Air Blank		0.000	20:18
Control Test		0.078	20:18
Air Blank		0.000	20:18
Diagnostics Check		OK	20:19

Data Download Date/Time: 11/18/2019 10:01

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

INSTRUMENT REGISTERED TO: HOLMES BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HURT,ALEXANDER,C
BREATH TEST OPERATOR AGENCY: Holmes Beach PD
ARRESTING OFFICER: HURT,ALEXANDER
ARRESTING OFFICER AGENCY: Holmes Beach PD
LAST AGENCY INSPECTION: 11/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175970
DRY GAS EXPIRY DATE: 08/02/2022
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: SABATO

DATE/TIME OF TEST: 11/24/2019 02:40:15

Diagnostics Check		OK	02:37	
Air Blank		0.000	02:37	
Control Test		0.081	02:38	
Air Blank		0.000	02:38	
Subject Sample	# 1		02:38	Subject Test Refused
Breath Volume		0.000	02:38	
Air Blank		0.000	02:39	
Control Test		0.081	02:39	
Air Blank		0.000	02:40	
Diagnostics Check		OK	02:40	

Data Download Date/Time: 12/16/2019 11:41

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006760

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUBEL,NEIL,F
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RUBEL,NEIL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: NAEGELE

DATE/TIME OF TEST: 11/28/2019 00:42:09

Diagnostics Check		OK	00:35
Air Blank		0.000	00:36
Control Test		0.080	00:36
Air Blank		0.000	00:36
Subject Sample	# 1	0.198	00:37
Breath Volume		1.824	00:37
Air Blank		0.000	00:38
Air Blank		0.000	00:39
Subject Sample	# 2	0.184	00:40
Breath Volume		1.621	00:40
Air Blank		0.000	00:41
Control Test		0.079	00:41
Air Blank		0.000	00:42
Diagnostics Check		OK	00:42

Data Download Date/Time: 12/12/2019 01:20

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006760

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUBEL,NEIL,F
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RUBEL,NEIL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/05/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: KELLEY

DATE/TIME OF TEST: 11/28/2019 03:00:17

Diagnostics Check		OK	02:52
Air Blank		0.000	02:53
Control Test		0.081	02:53
Air Blank		0.000	02:54
Subject Sample	# 1	0.175	02:55
Breath Volume		1.796	02:55
Air Blank		0.000	02:56
Air Blank		0.000	02:58
Subject Sample	# 2	0.170	02:58
Breath Volume		1.757	02:58
Air Blank		0.000	02:59
Control Test		0.080	02:59
Air Blank		0.000	03:00
Diagnostics Check		OK	03:00

Data Download Date/Time: 12/12/2019 01:20

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 10/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: NIELSEN

DATE/TIME OF TEST: 11/02/2019 00:30:29

Diagnostics Check		OK	00:24
Air Blank		0.000	00:24
Control Test		0.078	00:25
Air Blank		0.000	00:25
Subject Sample	# 1	0.136	00:26
Breath Volume		2.816	00:26
Air Blank		0.000	00:26
Air Blank		0.000	00:28
Subject Sample	# 2	0.135	00:28
Breath Volume		2.882	00:28
Air Blank		0.000	00:29
Control Test		0.077	00:29
Air Blank		0.000	00:30
Diagnostics Check		OK	00:30

Data Download Date/Time: 11/20/2019 20:01

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 10/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: GOOLSBY

DATE/TIME OF TEST: 11/02/2019 03:26:54

Diagnostics Check		OK	03:15
Air Blank		0.000	03:16
Control Test		0.077	03:16
Air Blank		0.000	03:16
Subject Sample	# 1	0.156	03:18
Breath Volume		1.292	03:18
Air Blank		0.000	03:18
Air Blank		0.000	03:20
Subject Sample	# 2	0.181	03:21
Breath Volume		1.937	03:21
Air Blank		0.000	03:22
Air Blank		0.000	03:24
Subject Sample	# 3	0.157	03:25
Breath Volume		1.250	03:25
Air Blank		0.000	03:25
Control Test		0.076	03:26
Air Blank		0.000	03:26
Diagnostics Check		OK	03:26

Data Download Date/Time: 11/20/2019 20:01

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 10/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: RUIZ

DATE/TIME OF TEST: 11/13/2019 00:07:38

Diagnostics Check		OK	00:01
Air Blank		0.000	00:01
Control Test		0.078	00:01
Air Blank		0.000	00:02
Subject Sample	# 1	0.164	00:03
Breath Volume		2.343	00:03
Air Blank		0.000	00:03
Air Blank		0.000	00:05
Subject Sample	# 2	0.174	00:06
Breath Volume		4.046	00:06
Air Blank		0.000	00:06
Control Test		0.077	00:07
Air Blank		0.000	00:07
Diagnostics Check		OK	00:07

Data Download Date/Time: 11/20/2019 20:01

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: LITTLE,D
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: HOLLEMAN

DATE/TIME OF TEST: 11/20/2019 23:57:24

Diagnostics Check		OK	23:50
Air Blank		0.000	23:51
Control Test		0.079	23:51
Air Blank		0.000	23:52
Subject Sample	# 1	0.126	23:52
Breath Volume		3.937	23:52
Air Blank		0.000	23:53
Air Blank		0.000	23:55
Subject Sample	# 2	0.122	23:55
Breath Volume		3.898	23:55
Air Blank		0.000	23:56
Control Test		0.078	23:56
Air Blank		0.000	23:57
Diagnostics Check		OK	23:57

Data Download Date/Time: 12/04/2019 19:28

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: KIRCHNER

DATE/TIME OF TEST: 11/21/2019 02:08:41

Diagnostics Check		OK	02:01
Air Blank		0.000	02:01
Control Test		0.079	02:02
Air Blank		0.000	02:02
Subject Sample	# 1	0.083	02:03
Breath Volume		3.394	02:03
Air Blank		0.000	02:04
Air Blank		0.000	02:06
Subject Sample	# 2	0.083	02:07
Breath Volume		3.058	02:07
Air Blank		0.000	02:07
Control Test		0.078	02:08
Air Blank		0.000	02:08
Diagnostics Check		OK	02:08

Data Download Date/Time: 12/04/2019 19:28

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: KEHOE

DATE/TIME OF TEST: 11/26/2019 22:58:18

Diagnostics Check		OK	22:50
Air Blank		0.000	22:51
Control Test		0.079	22:51
Air Blank		0.000	22:52
Subject Sample	# 1	0.206	22:53
Breath Volume		1.898	22:53
Air Blank		0.000	22:54
Air Blank		0.000	22:56
Subject Sample	# 2	0.188	22:56
Breath Volume		1.382	22:56
Air Blank		0.000	22:57
Control Test		0.078	22:57
Air Blank		0.000	22:58
Diagnostics Check		OK	22:58

Data Download Date/Time: 12/04/2019 19:28

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 10/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: FIX

DATE/TIME OF TEST: 11/01/2019 00:28:28

Diagnostics Check		OK	00:22
Air Blank		0.000	00:22
Control Test		0.081	00:22
Air Blank		0.000	00:23
Subject Sample	# 1	0.031	00:24
Breath Volume		1.511	00:24
Air Blank		0.000	00:24
Air Blank		0.000	00:26
Subject Sample	# 2	0.031	00:26
Breath Volume		2.375	00:26
Air Blank		0.000	00:27
Control Test		0.081	00:27
Air Blank		0.000	00:28
Diagnostics Check		OK	00:28

Data Download Date/Time: 11/20/2019 19:30

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 10/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: VARNUM

DATE/TIME OF TEST: 11/02/2019 02:40:42

Diagnostics Check		OK	02:34
Air Blank		0.000	02:34
Control Test		0.081	02:35
Air Blank		0.000	02:35
Subject Sample	# 1	0.109	02:36
Breath Volume		1.933	02:36
Air Blank		0.000	02:36
Air Blank		0.000	02:38
Subject Sample	# 2	0.105	02:39
Breath Volume		1.511	02:39
Air Blank		0.000	02:39
Control Test		0.082	02:40
Air Blank		0.000	02:40
Diagnostics Check		OK	02:40

Data Download Date/Time: 11/20/2019 19:30

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 10/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 11/07/2019 01:24:45

Diagnostics Check		OK	01:17
Air Blank		0.000	01:18
Control Test		0.082	01:18
Air Blank		0.000	01:19
Subject Sample	# 1	0.094	01:20
Breath Volume		1.964	01:20
Air Blank		0.000	01:20
Air Blank		0.000	01:22
Subject Sample	# 2	0.095	01:23
Breath Volume		2.050	01:23
Air Blank		0.000	01:23
Control Test		0.081	01:24
Air Blank		0.000	01:24
Diagnostics Check		OK	01:24

Data Download Date/Time: 11/20/2019 19:30

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: KHAIRY

DATE/TIME OF TEST: 11/29/2019 02:39:57

Diagnostics Check		OK	02:33
Air Blank		0.000	02:34
Control Test		0.082	02:34
Air Blank		0.000	02:34
Subject Sample	# 1	0.099	02:35
Breath Volume		1.589	02:35
Air Blank		0.000	02:36
Air Blank		0.000	02:37
Subject Sample	# 2	0.096	02:38
Breath Volume		1.718	02:38
Air Blank		0.000	02:39
Control Test		0.082	02:39
Air Blank		0.000	02:39
Diagnostics Check		OK	02:39

Data Download Date/Time: 12/04/2019 19:47

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006768

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREIRA,JOSE,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: PEREIRA,JOSE
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 18:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: LEBAKKEN

DATE/TIME OF TEST: 11/29/2019 19:06:38

Diagnostics Check		OK	18:56
Air Blank		0.000	18:56
Control Test		0.080	18:56
Air Blank		0.000	18:57
Subject Sample	# 1	0.341	18:58
Breath Volume		2.164	18:58
Air Blank		0.000	18:59
Air Blank		0.000	19:01
Subject Sample	# 2	0.383	19:01
Breath Volume		2.019	19:01
Air Blank		0.000	19:02
Air Blank		0.000	19:04
Subject Sample	# 3	0.377	19:04
Breath Volume		2.046	19:04
Air Blank		0.000	19:05
Control Test		0.078	19:06
Air Blank		0.000	19:06
Diagnostics Check		OK	19:06

Data Download Date/Time: 12/12/2019 10:47

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: DENBOW,M
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 10/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 11/01/2019 00:56:45

Diagnostics Check		OK	00:49
Air Blank		0.000	00:50
Control Test		0.078	00:50
Air Blank		0.000	00:51
Subject Sample	# 1	0.207	00:51
Breath Volume		2.398	00:51
Air Blank		0.000	00:52
Air Blank		0.000	00:54
Subject Sample	# 2	0.203	00:55
Breath Volume		2.257	00:55
Air Blank		0.000	00:55
Control Test		0.077	00:56
Air Blank		0.000	00:56
Diagnostics Check		OK	00:56

Data Download Date/Time: 11/15/2019 12:12

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 10/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:52

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 11/01/2019 05:26:18

Diagnostics Check		OK	05:17
Air Blank		0.000	05:18
Control Test		0.077	05:18
Air Blank		0.000	05:19
Subject Sample	# 1	0.126	05:21
Breath Volume		1.792	05:21
Air Blank		0.000	05:22
Air Blank		0.000	05:23
Subject Sample	# 2	0.122	05:24
Breath Volume		1.753	05:24
Air Blank		0.000	05:25
Control Test		0.077	05:25
Air Blank		0.000	05:26
Diagnostics Check		OK	05:26

Data Download Date/Time: 11/15/2019 12:12

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 10/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 06:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:29

SUBJECT'S LAST NAME: MEYERSON

DATE/TIME OF TEST: 11/02/2019 08:03:25

Diagnostics Check		OK	07:56
Air Blank		0.000	07:57
Control Test		0.077	07:57
Air Blank		0.000	07:58
Subject Sample	# 1	0.101	07:58
Breath Volume		1.976	07:58
Air Blank		0.000	07:59
Air Blank		0.000	08:01
Subject Sample	# 2	0.103	08:01
Breath Volume		1.863	08:01
Air Blank		0.000	08:02
Control Test		0.077	08:02
Air Blank		0.000	08:03
Diagnostics Check		OK	08:03

Data Download Date/Time: 11/15/2019 12:12

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 10/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: VILLALOBOS

DATE/TIME OF TEST: 11/03/2019 01:32:16

Diagnostics Check		OK	01:29	
Air Blank		0.000	01:29	
Control Test		0.078	01:29	
Air Blank		0.000	01:30	
Subject Sample	# 1		01:30	Subject Test Refused
Breath Volume		0.000	01:30	
Air Blank		0.000	01:31	
Control Test		0.077	01:31	
Air Blank		0.000	01:32	
Diagnostics Check		OK	01:32	

Data Download Date/Time: 11/15/2019 12:12

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: FERRO,L
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 10/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:06

SUBJECT'S LAST NAME: GARCIA GUTIERREZ

DATE/TIME OF TEST: 11/08/2019 22:44:31

Diagnostics Check		OK	22:41	
Air Blank		0.000	22:42	
Control Test		0.078	22:42	
Air Blank		0.000	22:42	
Subject Sample	# 1		22:43	Subject Test Refused
Breath Volume		0.000	22:43	
Air Blank		0.000	22:43	
Control Test		0.076	22:43	
Air Blank		0.000	22:44	
Diagnostics Check		OK	22:44	

Data Download Date/Time: 11/15/2019 12:12

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: MCGRATH,R
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 10/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: FLORES

DATE/TIME OF TEST: 11/09/2019 02:30:30

Diagnostics Check		OK	02:21
Air Blank		0.000	02:21
Control Test		0.078	02:21
Air Blank		0.000	02:22
Subject Sample	# 1	0.114	02:23
Breath Volume		1.441	02:23
Air Blank		0.000	02:23
Air Blank		0.000	02:25
Subject Sample	# 2	0.127	02:28
Breath Volume		1.289	02:28
Air Blank		0.000	02:29
Control Test		0.079	02:29
Air Blank		0.000	02:30
Diagnostics Check		OK	02:30

Data Download Date/Time: 11/15/2019 12:12

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 10/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 06:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:35

SUBJECT'S LAST NAME: VILLAR MORAN

DATE/TIME OF TEST: 11/10/2019 08:13:17

Diagnostics Check		OK	08:10	
Air Blank		0.000	08:10	
Control Test		0.078	08:10	
Air Blank		0.000	08:11	
Subject Sample	# 1		08:11	Subject Test Refused
Breath Volume		0.000	08:11	
Air Blank		0.000	08:12	
Control Test		0.078	08:12	
Air Blank		0.000	08:13	
Diagnostics Check		OK	08:13	

Data Download Date/Time: 11/15/2019 12:12

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: RAWLUK,S
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 10/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 07:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:30

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 11/10/2019 08:30:31

Diagnostics Check		OK	08:23
Air Blank		0.000	08:24
Control Test		0.078	08:24
Air Blank		0.000	08:25
Subject Sample	# 1	0.169	08:25
Breath Volume		2.117	08:25
Air Blank		0.000	08:26
Air Blank		0.000	08:28
Subject Sample	# 2	0.170	08:28
Breath Volume		2.621	08:28
Air Blank		0.000	08:29
Control Test		0.078	08:29
Air Blank		0.000	08:30
Diagnostics Check		OK	08:30

Data Download Date/Time: 11/15/2019 12:12

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: FRANCO,G
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: RICHARD

DATE/TIME OF TEST: 11/16/2019 04:12:06

Diagnostics Check		OK	04:08	
Air Blank		0.000	04:09	
Control Test		0.079	04:09	
Air Blank		0.000	04:10	
Subject Sample	# 1		04:10	Subject Test Refused
Breath Volume		0.000	04:10	
Air Blank		0.000	04:11	
Control Test		0.078	04:11	
Air Blank		0.000	04:12	
Diagnostics Check		OK	04:12	

Data Download Date/Time: 12/27/2019 11:24

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: QUINTANA,J
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: VAZQUEZ HERMOSA

DATE/TIME OF TEST: 11/17/2019 02:39:52

Diagnostics Check		OK	02:33
Air Blank		0.000	02:33
Control Test		0.079	02:33
Air Blank		0.000	02:34
Subject Sample	# 1	0.254	02:35
Breath Volume		1.750	02:35
Air Blank		0.000	02:35
Air Blank		0.000	02:37
Subject Sample	# 2	0.247	02:38
Breath Volume		1.523	02:38
Air Blank		0.000	02:38
Control Test		0.079	02:39
Air Blank		0.000	02:39
Diagnostics Check		OK	02:39

Data Download Date/Time: 12/27/2019 11:24

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: BELLEVY

DATE/TIME OF TEST: 11/17/2019 05:02:29

Diagnostics Check		OK	04:59	
Air Blank		0.000	04:59	
Control Test		0.080	05:00	
Air Blank		0.000	05:00	
Subject Sample	# 1		05:01	Subject Test Refused
Breath Volume		0.000	05:01	
Air Blank		0.000	05:01	
Control Test		0.077	05:01	
Air Blank		0.000	05:02	
Diagnostics Check		OK	05:02	

Data Download Date/Time: 12/27/2019 11:24

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: QUINTANA,J
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:57

SUBJECT'S LAST NAME: LOSME

DATE/TIME OF TEST: 11/17/2019 06:36:00

Diagnostics Check		OK	06:29
Air Blank		0.000	06:29
Control Test		0.080	06:29
Air Blank		0.000	06:30
Subject Sample	# 1	0.180	06:31
Breath Volume		2.828	06:31
Air Blank		0.000	06:31
Air Blank		0.000	06:33
Subject Sample	# 2	0.183	06:34
Breath Volume		2.687	06:34
Air Blank		0.000	06:35
Control Test		0.080	06:35
Air Blank		0.000	06:35
Diagnostics Check		OK	06:36

Data Download Date/Time: 12/27/2019 11:24

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:03

SUBJECT'S LAST NAME: LLEONART

DATE/TIME OF TEST: 11/23/2019 05:36:31

Diagnostics Check		OK	05:33	
Air Blank		0.000	05:34	
Control Test		0.078	05:34	
Air Blank		0.000	05:34	
Subject Sample	# 1		05:35	Subject Test Refused
Breath Volume		0.000	05:35	
Air Blank		0.000	05:35	
Control Test		0.077	05:35	
Air Blank		0.000	05:36	
Diagnostics Check		OK	05:36	

Data Download Date/Time: 12/27/2019 11:24

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: QUINTANA,J
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:35

SUBJECT'S LAST NAME: V622540800570

DATE/TIME OF TEST: 11/23/2019 06:07:50

Diagnostics Check		OK	06:04	
Air Blank		0.000	06:05	
Control Test		0.079	06:05	
Air Blank		0.000	06:06	
Subject Sample	# 1		06:06	Subject Test Refused
Breath Volume		0.000	06:06	
Air Blank		0.000	06:06	
Control Test		0.077	06:07	
Air Blank		0.000	06:07	
Diagnostics Check		OK	06:07	

Data Download Date/Time: 12/27/2019 11:24

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: FRANCO,G
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 11/24/2019 01:06:15

Diagnostics Check		OK	00:59
Air Blank		0.000	00:59
Control Test		0.079	01:00
Air Blank		0.000	01:00
Subject Sample	# 1	0.072	01:01
Breath Volume		1.675	01:01
Air Blank		0.000	01:01
Air Blank		0.000	01:03
Subject Sample	# 2	0.068	01:04
Breath Volume		1.464	01:04
Air Blank		0.000	01:05
Control Test		0.080	01:05
Air Blank		0.000	01:06
Diagnostics Check		OK	01:06

Data Download Date/Time: 12/27/2019 11:24

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: QUINTANA,J
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:59

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 11/24/2019 06:34:19

Diagnostics Check		OK	06:31	
Air Blank		0.000	06:31	
Control Test		0.080	06:32	
Air Blank		0.000	06:32	
Subject Sample	# 1		06:32	Subject Test Refused
Breath Volume		0.000	06:32	
Air Blank		0.000	06:33	
Control Test		0.078	06:33	
Air Blank		0.000	06:34	
Diagnostics Check		OK	06:34	

Data Download Date/Time: 12/27/2019 11:24

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: SORIANO

DATE/TIME OF TEST: 11/28/2019 02:05:33

Diagnostics Check		OK	01:58
Air Blank		0.000	01:59
Control Test		0.079	01:59
Air Blank		0.000	01:59
Subject Sample	# 1	0.136	02:00
Breath Volume		1.828	02:00
Air Blank		0.000	02:01
Air Blank		0.000	02:03
Subject Sample	# 2	0.134	02:03
Breath Volume		1.722	02:03
Air Blank		0.000	02:04
Control Test		0.079	02:04
Air Blank		0.000	02:05
Diagnostics Check		OK	02:05

Data Download Date/Time: 12/27/2019 11:24

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: DELGADO,M
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: DAUPHIN

DATE/TIME OF TEST: 11/29/2019 00:50:36

Diagnostics Check		OK	00:43
Air Blank		0.000	00:43
Control Test		0.079	00:44
Air Blank		0.000	00:44
Subject Sample	# 1	0.238	00:45
Breath Volume		1.222	00:45
Air Blank		0.000	00:46
Air Blank		0.000	00:48
Subject Sample	# 2	0.252	00:49
Breath Volume		1.664	00:49
Air Blank		0.000	00:49
Control Test		0.079	00:50
Air Blank		0.000	00:50
Diagnostics Check		OK	00:50

Data Download Date/Time: 12/27/2019 11:24

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: COLINDRES

DATE/TIME OF TEST: 11/29/2019 04:29:57

Diagnostics Check		OK	04:22
Air Blank		0.000	04:23
Control Test		0.078	04:23
Air Blank		0.000	04:23
Subject Sample	# 1	0.157	04:25
Breath Volume		2.039	04:25
Air Blank		0.000	04:25
Air Blank		0.000	04:27
Subject Sample	# 2	0.160	04:28
Breath Volume		2.128	04:28
Air Blank		0.000	04:29
Control Test		0.080	04:29
Air Blank		0.000	04:29
Diagnostics Check		OK	04:29

Data Download Date/Time: 12/27/2019 11:24

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: LARROSA,C
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 06:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:14

SUBJECT'S LAST NAME: RICHARDS

DATE/TIME OF TEST: 11/29/2019 07:48:07

Diagnostics Check		OK	07:45	
Air Blank		0.000	07:45	
Control Test		0.079	07:45	
Air Blank		0.000	07:46	
Subject Sample	# 1		07:46	Subject Test Refused
Breath Volume		0.000	07:46	
Air Blank		0.000	07:47	
Control Test		0.078	07:47	
Air Blank		0.000	07:48	
Diagnostics Check		OK	07:48	

Data Download Date/Time: 12/27/2019 11:24

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: QUINTANA,J
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 11/30/2019 01:36:33

Diagnostics Check		OK	01:29
Air Blank		0.000	01:29
Control Test		0.078	01:30
Air Blank		0.000	01:30
Subject Sample	# 1	0.162	01:31
Breath Volume		1.601	01:31
Air Blank		0.000	01:32
Air Blank		0.000	01:34
Subject Sample	# 2	0.161	01:35
Breath Volume		1.847	01:35
Air Blank		0.000	01:35
Control Test		0.078	01:35
Air Blank		0.000	01:36
Diagnostics Check		OK	01:36

Data Download Date/Time: 12/27/2019 11:24

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:01

SUBJECT'S LAST NAME: ORJUELA

DATE/TIME OF TEST: 11/30/2019 05:32:51

Diagnostics Check		OK	05:26
Air Blank		0.000	05:26
Control Test		0.079	05:27
Air Blank		0.000	05:27
Subject Sample	# 1	0.210	05:28
Breath Volume		2.507	05:28
Air Blank		0.000	05:29
Air Blank		0.000	05:30
Subject Sample	# 2	0.219	05:31
Breath Volume		2.082	05:31
Air Blank		0.000	05:31
Control Test		0.079	05:32
Air Blank		0.000	05:32
Diagnostics Check		OK	05:32

Data Download Date/Time: 12/27/2019 11:24

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DERIET,RAFAEL
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: BITTER

DATE/TIME OF TEST: 11/14/2019 01:51:35

Diagnostics Check		OK	01:48	
Air Blank		0.000	01:48	
Control Test		0.079	01:48	
Air Blank		0.000	01:49	
Subject Sample	# 1		01:50	Interferent Detect
Breath Volume		2.652	01:50	
Air Blank		0.000	01:50	
Control Test		0.080	01:50	
Air Blank		0.000	01:51	
Diagnostics Check		OK	01:51	

Data Download Date/Time: 12/05/2019 12:14

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: PIERRE

DATE/TIME OF TEST: 11/16/2019 02:12:14

Diagnostics Check		OK	02:05
Air Blank		0.000	02:06
Control Test		0.080	02:06
Air Blank		0.000	02:07
Subject Sample	# 1	0.172	02:07
Breath Volume		2.093	02:07
Air Blank		0.000	02:08
Air Blank		0.000	02:10
Subject Sample	# 2	0.173	02:10
Breath Volume		2.683	02:10
Air Blank		0.000	02:11
Control Test		0.080	02:11
Air Blank		0.000	02:12
Diagnostics Check		OK	02:12

Data Download Date/Time: 12/05/2019 12:14

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LALANE,LYNN
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 07:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:19

SUBJECT'S LAST NAME: MEJIA

DATE/TIME OF TEST: 11/16/2019 08:47:06

Diagnostics Check		OK	08:40
Air Blank		0.000	08:41
Control Test		0.080	08:41
Air Blank		0.000	08:42
Subject Sample	# 1	0.131	08:42
Breath Volume		1.656	08:42
Air Blank		0.000	08:43
Air Blank		0.000	08:45
Subject Sample	# 2	0.133	08:45
Breath Volume		3.039	08:45
Air Blank		0.000	08:46
Control Test		0.078	08:46
Air Blank		0.000	08:47
Diagnostics Check		OK	08:47

Data Download Date/Time: 12/05/2019 12:14

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: PARRISH,MATTHEW
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 11:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:08

SUBJECT'S LAST NAME: ZACARIAS ROJAS

DATE/TIME OF TEST: 11/16/2019 12:40:25

Diagnostics Check		OK	12:33
Air Blank		0.000	12:33
Control Test		0.081	12:33
Air Blank		0.000	12:34
Subject Sample	# 1	0.204	12:35
Breath Volume		1.636	12:35
Air Blank		0.000	12:36
Air Blank		0.000	12:38
Subject Sample	# 2	0.194	12:38
Breath Volume		1.968	12:38
Air Blank		0.000	12:39
Control Test		0.077	12:39
Air Blank		0.000	12:40
Diagnostics Check		OK	12:40

Data Download Date/Time: 12/05/2019 12:14

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAPP,JOSHUA,R
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KEATING,K
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 18:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:48

SUBJECT'S LAST NAME: ELEJALDE

DATE/TIME OF TEST: 11/16/2019 19:21:31

Diagnostics Check		OK	19:14
Air Blank		0.000	19:15
Control Test		0.081	19:15
Air Blank		0.000	19:16
Subject Sample	# 1	0.187	19:17
Breath Volume		2.250	19:17
Air Blank		0.000	19:17
Air Blank		0.000	19:19
Subject Sample	# 2	0.187	19:19
Breath Volume		1.488	19:19
Air Blank		0.000	19:20
Control Test		0.078	19:20
Air Blank		0.000	19:21
Diagnostics Check		OK	19:21

Data Download Date/Time: 12/05/2019 12:14

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SGT. BOHM,JOHN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: SUCKRAJH

DATE/TIME OF TEST: 11/17/2019 02:39:06

Diagnostics Check		OK	02:26	
Air Blank		0.000	02:27	
Control Test		0.080	02:27	
Air Blank		0.000	02:27	
Subject Sample	# 1	0.138	02:31	Volume Not Met
Breath Volume		0.582	02:31	
Air Blank		0.000	02:31	
Air Blank		0.000	02:33	
Subject Sample	# 2	0.172	02:34	
Breath Volume		2.128	02:34	
Air Blank		0.000	02:34	
Air Blank		0.000	02:36	
Subject Sample	# 3	0.184	02:37	
Breath Volume		1.828	02:37	
Air Blank		0.000	02:38	
Control Test		0.079	02:38	
Air Blank		0.000	02:39	
Diagnostics Check		OK	02:39	

Data Download Date/Time: 12/05/2019 12:14

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SGT. BOHM,JONATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:36

SUBJECT'S LAST NAME: BIRNBAUM

DATE/TIME OF TEST: 11/17/2019 05:05:24

Diagnostics Check		OK	04:58
Air Blank		0.000	04:58
Control Test		0.079	04:58
Air Blank		0.000	04:59
Subject Sample	# 1	0.124	04:59
Breath Volume		2.511	04:59
Air Blank		0.000	05:00
Air Blank		0.000	05:02
Subject Sample	# 2	0.120	05:03
Breath Volume		2.839	05:03
Air Blank		0.000	05:04
Control Test		0.079	05:04
Air Blank		0.000	05:05
Diagnostics Check		OK	05:05

Data Download Date/Time: 12/05/2019 12:14

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: O'DOR,ROBERT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 05:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:51

SUBJECT'S LAST NAME: NESTOR

DATE/TIME OF TEST: 11/17/2019 06:20:56

Diagnostics Check		OK	06:14
Air Blank		0.000	06:15
Control Test		0.079	06:15
Air Blank		0.000	06:15
Subject Sample	# 1	0.179	06:16
Breath Volume		1.273	06:16
Air Blank		0.000	06:17
Air Blank		0.000	06:18
Subject Sample	# 2	0.177	06:19
Breath Volume		1.222	06:19
Air Blank		0.000	06:20
Control Test		0.078	06:20
Air Blank		0.000	06:20
Diagnostics Check		OK	06:20

Data Download Date/Time: 12/05/2019 12:14

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BOONE,MICHAEL
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: NAIT

DATE/TIME OF TEST: 11/18/2019 01:00:25

Diagnostics Check		OK	00:53
Air Blank		0.000	00:54
Control Test		0.080	00:54
Air Blank		0.000	00:54
Subject Sample	# 1	0.192	00:55
Breath Volume		1.218	00:55
Air Blank		0.000	00:56
Air Blank		0.000	00:57
Subject Sample	# 2	0.197	00:58
Breath Volume		1.292	00:58
Air Blank		0.000	00:59
Control Test		0.079	00:59
Air Blank		0.000	01:00
Diagnostics Check		OK	01:00

Data Download Date/Time: 12/05/2019 12:14

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: PAUL,STIVENSON
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: BARTON

DATE/TIME OF TEST: 11/21/2019 23:36:48

Diagnostics Check		OK	23:29
Air Blank		0.000	23:30
Control Test		0.080	23:30
Air Blank		0.000	23:30
Subject Sample	# 1	0.126	23:31
Breath Volume		2.636	23:31
Air Blank		0.000	23:32
Air Blank		0.000	23:34
Subject Sample	# 2	0.128	23:35
Breath Volume		2.074	23:35
Air Blank		0.000	23:35
Control Test		0.080	23:36
Air Blank		0.000	23:36
Diagnostics Check		OK	23:36

Data Download Date/Time: 12/05/2019 12:14

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: O'DOR,ROBERT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: RENTZ

DATE/TIME OF TEST: 11/22/2019 03:39:43

Diagnostics Check		OK	03:32
Air Blank		0.000	03:33
Control Test		0.079	03:33
Air Blank		0.000	03:34
Subject Sample	# 1	0.235	03:34
Breath Volume		1.964	03:34
Air Blank		0.000	03:35
Air Blank		0.000	03:36
Subject Sample	# 2	0.227	03:37
Breath Volume		2.925	03:37
Air Blank		0.000	03:38
Control Test		0.079	03:39
Air Blank		0.000	03:39
Diagnostics Check		OK	03:39

Data Download Date/Time: 12/05/2019 12:14

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 04:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:09

SUBJECT'S LAST NAME: ROLLINS

DATE/TIME OF TEST: 11/22/2019 05:44:39

Diagnostics Check		OK	05:37
Air Blank		0.000	05:38
Control Test		0.079	05:38
Air Blank		0.000	05:39
Subject Sample	# 1	0.093	05:39
Breath Volume		2.105	05:39
Air Blank		0.000	05:40
Air Blank		0.000	05:42
Subject Sample	# 2	0.092	05:43
Breath Volume		1.855	05:43
Air Blank		0.000	05:43
Control Test		0.079	05:44
Air Blank		0.000	05:44
Diagnostics Check		OK	05:44

Data Download Date/Time: 12/05/2019 12:14

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HOLLOWAY,DEVARIOUS
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: CUNNINGHAM

DATE/TIME OF TEST: 11/23/2019 01:10:55

Diagnostics Check		OK	00:54	
Air Blank		0.000	00:55	
Control Test		0.079	00:55	
Air Blank		0.000	00:55	
Subject Sample	# 1	0.165	00:59	Volume Not Met
Breath Volume		0.464	00:59	
Air Blank		0.000	00:59	
Air Blank		0.000	01:01	
Subject Sample	# 2	0.164	01:03	
Breath Volume		1.195	01:03	
Air Blank		0.000	01:04	
Air Blank		0.000	01:06	
Subject Sample	# 3	0.109	01:09	Volume Not Met
Breath Volume		0.371	01:09	
Air Blank		0.000	01:09	
Control Test		0.078	01:10	
Air Blank		0.000	01:10	
Diagnostics Check		OK	01:10	

Data Download Date/Time: 12/05/2019 12:14

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HOLLOWAY,DEVARIOUS
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: CUNNINGHAM

DATE/TIME OF TEST: 11/23/2019 01:20:29

Diagnostics Check		OK	01:13
Air Blank		0.000	01:14
Control Test		0.077	01:14
Air Blank		0.000	01:15
Subject Sample	# 1	0.150	01:15
Breath Volume		1.457	01:15
Air Blank		0.000	01:16
Air Blank		0.000	01:18
Subject Sample	# 2	0.155	01:18
Breath Volume		1.847	01:18
Air Blank		0.000	01:19
Control Test		0.078	01:19
Air Blank		0.000	01:20
Diagnostics Check		OK	01:20

Data Download Date/Time: 12/05/2019 12:14

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ULLOA,ADRIAN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: RICHET

DATE/TIME OF TEST: 11/23/2019 03:36:19

Diagnostics Check		OK	03:30
Air Blank		0.000	03:30
Control Test		0.079	03:30
Air Blank		0.000	03:31
Subject Sample	# 1	0.183	03:31
Breath Volume		1.191	03:31
Air Blank		0.000	03:32
Air Blank		0.000	03:34
Subject Sample	# 2	0.181	03:34
Breath Volume		1.343	03:34
Air Blank		0.000	03:35
Control Test		0.078	03:35
Air Blank		0.000	03:36
Diagnostics Check		OK	03:36

Data Download Date/Time: 12/05/2019 12:14

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 20:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: GUTIERREZ REYES

DATE/TIME OF TEST: 11/24/2019 21:47:35

Diagnostics Check		OK	21:41
Air Blank		0.000	21:41
Control Test		0.079	21:42
Air Blank		0.000	21:42
Subject Sample	# 1	0.246	21:43
Breath Volume		1.289	21:43
Air Blank		0.000	21:43
Air Blank		0.000	21:45
Subject Sample	# 2	0.241	21:45
Breath Volume		1.242	21:45
Air Blank		0.000	21:46
Control Test		0.079	21:46
Air Blank		0.000	21:47
Diagnostics Check		OK	21:47

Data Download Date/Time: 12/05/2019 12:14

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: O'DOR,ROBERT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:24

SUBJECT'S LAST NAME: FRANCISCO

DATE/TIME OF TEST: 11/24/2019 22:01:19

Diagnostics Check		OK	21:54
Air Blank		0.000	21:55
Control Test		0.078	21:55
Air Blank		0.000	21:55
Subject Sample	# 1	0.141	21:56
Breath Volume		1.890	21:56
Air Blank		0.000	21:57
Air Blank		0.000	21:59
Subject Sample	# 2	0.138	21:59
Breath Volume		1.652	21:59
Air Blank		0.000	22:00
Control Test		0.078	22:00
Air Blank		0.000	22:01
Diagnostics Check		OK	22:01

Data Download Date/Time: 12/05/2019 12:14

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HORNE,PHILLIP
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 15:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:26

SUBJECT'S LAST NAME: CHERISMA

DATE/TIME OF TEST: 11/27/2019 17:07:12

Diagnostics Check		OK	16:55	
Air Blank		0.000	16:56	
Control Test		0.080	16:56	
Air Blank		0.000	16:56	
Subject Sample	# 1	0.296	16:57	
Breath Volume		1.191	16:57	
Air Blank		0.000	16:58	
Air Blank		0.000	16:59	
Subject Sample	# 2	0.271	17:02	
Breath Volume		1.503	17:02	
Air Blank		0.000	17:03	
Air Blank		0.000	17:05	
Subject Sample	# 3	0.291	17:05	
Breath Volume		1.468	17:05	
Air Blank		0.000	17:06	
Control Test		0.074	17:06	Control Outside Tolerance
Air Blank		0.000	17:07	

Data Download Date/Time: 12/05/2019 12:14

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HORNE,PHILLIP
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 15:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:26

SUBJECT'S LAST NAME: CHERISMA

DATE/TIME OF TEST: 11/27/2019 17:15:58

Diagnostics Check		OK	17:08
Air Blank		0.000	17:09
Control Test		0.077	17:09
Air Blank		0.000	17:10
Subject Sample	# 1	0.299	17:10
Breath Volume		1.828	17:10
Air Blank		0.000	17:11
Air Blank		0.000	17:13
Subject Sample	# 2	0.298	17:14
Breath Volume		2.023	17:14
Air Blank		0.000	17:15
Control Test		0.078	17:15
Air Blank		0.000	17:15
Diagnostics Check		OK	17:15

Data Download Date/Time: 12/05/2019 12:14

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: PRESUME,JEHUD
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: GORDON

DATE/TIME OF TEST: 11/29/2019 02:40:04

Diagnostics Check		OK	02:33
Air Blank		0.000	02:33
Control Test		0.079	02:34
Air Blank		0.000	02:34
Subject Sample	# 1	0.127	02:35
Breath Volume		3.660	02:35
Air Blank		0.000	02:36
Air Blank		0.000	02:37
Subject Sample	# 2	0.124	02:38
Breath Volume		3.117	02:38
Air Blank		0.000	02:39
Control Test		0.078	02:39
Air Blank		0.000	02:39
Diagnostics Check		OK	02:40

Data Download Date/Time: 12/05/2019 12:14

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 19:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:17

SUBJECT'S LAST NAME: MEDRANO

DATE/TIME OF TEST: 11/29/2019 20:45:50

Diagnostics Check		OK	20:38
Air Blank		0.000	20:39
Control Test		0.079	20:39
Air Blank		0.000	20:39
Subject Sample	# 1	0.144	20:41
Breath Volume		1.777	20:41
Air Blank		0.000	20:41
Air Blank		0.000	20:43
Subject Sample	# 2	0.142	20:44
Breath Volume		1.968	20:44
Air Blank		0.000	20:44
Control Test		0.078	20:45
Air Blank		0.000	20:45
Diagnostics Check		OK	20:45

Data Download Date/Time: 12/05/2019 12:14

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LUNA,KRISTINA
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 20:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:08

SUBJECT'S LAST NAME: CAREY

DATE/TIME OF TEST: 11/29/2019 21:43:31

Diagnostics Check		OK	21:36
Air Blank		0.000	21:37
Control Test		0.079	21:37
Air Blank		0.000	21:38
Subject Sample	# 1	0.007	21:39
Breath Volume		1.203	21:39
Air Blank		0.000	21:39
Air Blank		0.000	21:41
Subject Sample	# 2	0.000	21:42
Breath Volume		1.320	21:42
Air Blank		0.000	21:42
Control Test		0.078	21:42
Air Blank		0.000	21:43
Diagnostics Check		OK	21:43

Data Download Date/Time: 12/05/2019 12:14

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DENNIS,CALVIN
ARRESTING OFFICER AGENCY: Seminole Dept of LE
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 19:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: JUMPER

DATE/TIME OF TEST: 11/30/2019 21:15:03

Diagnostics Check		OK	21:07
Air Blank		0.000	21:08
Control Test		0.079	21:08
Air Blank		0.000	21:09
Subject Sample	# 1	0.204	21:10
Breath Volume		1.441	21:10
Air Blank		0.000	21:11
Air Blank		0.000	21:12
Subject Sample	# 2	0.192	21:13
Breath Volume		1.101	21:13
Air Blank		0.000	21:14
Control Test		0.077	21:14
Air Blank		0.000	21:14
Diagnostics Check		OK	21:15

Data Download Date/Time: 12/05/2019 12:14

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LARKIN,ARIJI
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: DENIS JR

DATE/TIME OF TEST: 11/01/2019 01:11:03

Diagnostics Check		OK	01:02
Air Blank		0.000	01:03
Control Test		0.080	01:03
Air Blank		0.000	01:04
Subject Sample	# 1	0.074	01:05
Breath Volume		1.136	01:05
Air Blank		0.000	01:06
Air Blank		0.000	01:07
Subject Sample	# 2	0.080	01:09
Breath Volume		1.816	01:09
Air Blank		0.000	01:10
Control Test		0.080	01:10
Air Blank		0.000	01:10
Diagnostics Check		OK	01:11

Data Download Date/Time: 11/13/2019 10:58

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: STONER,NATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: WATTS

DATE/TIME OF TEST: 11/01/2019 01:40:50

Diagnostics Check		OK	01:34
Air Blank		0.000	01:34
Control Test		0.079	01:35
Air Blank		0.000	01:35
Subject Sample	# 1	0.105	01:36
Breath Volume		4.417	01:36
Air Blank		0.000	01:36
Air Blank		0.000	01:38
Subject Sample	# 2	0.102	01:39
Breath Volume		2.917	01:39
Air Blank		0.000	01:39
Control Test		0.080	01:40
Air Blank		0.000	01:40
Diagnostics Check		OK	01:40

Data Download Date/Time: 11/13/2019 10:58

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 11/01/2019 02:34:02

Diagnostics Check		OK	02:25
Air Blank		0.000	02:25
Control Test		0.081	02:26
Air Blank		0.000	02:26
Subject Sample	# 1	0.151	02:29
Breath Volume		1.910	02:29
Air Blank		0.000	02:30
Air Blank		0.000	02:31
Subject Sample	# 2	0.158	02:32
Breath Volume		1.570	02:32
Air Blank		0.000	02:33
Control Test		0.079	02:33
Air Blank		0.000	02:33
Diagnostics Check		OK	02:34

Data Download Date/Time: 11/13/2019 10:58

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: STONER,NATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: PREVAL

DATE/TIME OF TEST: 11/01/2019 03:22:17

Diagnostics Check		OK	03:15
Air Blank		0.000	03:16
Control Test		0.080	03:16
Air Blank		0.000	03:16
Subject Sample	# 1	0.199	03:17
Breath Volume		2.324	03:17
Air Blank		0.000	03:18
Air Blank		0.000	03:20
Subject Sample	# 2	0.202	03:20
Breath Volume		2.121	03:20
Air Blank		0.000	03:21
Control Test		0.079	03:21
Air Blank		0.000	03:22
Diagnostics Check		OK	03:22

Data Download Date/Time: 11/13/2019 10:58

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FARQUHARSON,KEVIN
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: PEREZ CALDERON

DATE/TIME OF TEST: 11/01/2019 04:44:50

Diagnostics Check		OK	04:37
Air Blank		0.000	04:37
Control Test		0.080	04:38
Air Blank		0.000	04:38
Subject Sample	# 1	0.198	04:39
Breath Volume		2.164	04:39
Air Blank		0.000	04:40
Air Blank		0.000	04:42
Subject Sample	# 2	0.200	04:43
Breath Volume		3.023	04:43
Air Blank		0.000	04:43
Control Test		0.079	04:44
Air Blank		0.000	04:44
Diagnostics Check		OK	04:44

Data Download Date/Time: 11/13/2019 10:58

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: STACHURA,HAYDEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: DELEON

DATE/TIME OF TEST: 11/02/2019 01:40:02

Diagnostics Check		OK	01:32
Air Blank		0.000	01:33
Control Test		0.080	01:33
Air Blank		0.000	01:34
Subject Sample	# 1	0.114	01:35
Breath Volume		1.941	01:35
Air Blank		0.000	01:35
Air Blank		0.000	01:37
Subject Sample	# 2	0.112	01:38
Breath Volume		1.554	01:38
Air Blank		0.000	01:39
Control Test		0.081	01:39
Air Blank		0.000	01:39
Diagnostics Check		OK	01:40

Data Download Date/Time: 11/13/2019 10:58

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ROJAS,RANDY
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: IPARRAGUIRRE

DATE/TIME OF TEST: 11/02/2019 03:29:57

Diagnostics Check		OK	03:22
Air Blank		0.000	03:23
Control Test		0.080	03:23
Air Blank		0.000	03:24
Subject Sample	# 1	0.047	03:25
Breath Volume		2.843	03:25
Air Blank		0.000	03:25
Air Blank		0.000	03:27
Subject Sample	# 2	0.047	03:28
Breath Volume		2.175	03:28
Air Blank		0.000	03:29
Control Test		0.080	03:29
Air Blank		0.000	03:29
Diagnostics Check		OK	03:29

Data Download Date/Time: 11/13/2019 10:58

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: NEWTON

DATE/TIME OF TEST: 11/03/2019 00:48:11

Diagnostics Check		OK	00:40
Air Blank		0.000	00:40
Control Test		0.080	00:41
Air Blank		0.000	00:41
Subject Sample	# 1	0.248	00:42
Breath Volume		1.394	00:42
Air Blank		0.000	00:43
Air Blank		0.000	00:44
Subject Sample	# 2	0.233	00:46
Breath Volume		1.437	00:46
Air Blank		0.000	00:47
Control Test		0.078	00:47
Air Blank		0.000	00:48
Diagnostics Check		OK	00:48

Data Download Date/Time: 11/13/2019 10:58

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: STACHURA,HAYDEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: BRUTUS

DATE/TIME OF TEST: 11/03/2019 02:44:43

Diagnostics Check		OK	02:38
Air Blank		0.000	02:38
Control Test		0.079	02:39
Air Blank		0.000	02:39
Subject Sample	# 1	0.130	02:40
Breath Volume		2.484	02:40
Air Blank		0.000	02:40
Air Blank		0.000	02:42
Subject Sample	# 2	0.124	02:43
Breath Volume		1.257	02:43
Air Blank		0.000	02:43
Control Test		0.079	02:44
Air Blank		0.000	02:44
Diagnostics Check		OK	02:44

Data Download Date/Time: 11/13/2019 10:58

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: O'DOR,ROBERT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: PEREIRA

DATE/TIME OF TEST: 11/03/2019 03:41:02

Diagnostics Check		OK	03:26	
Air Blank		0.000	03:26	
Control Test		0.079	03:27	
Air Blank		0.000	03:27	
Subject Sample	# 1		03:30	No Sample Provided
Breath Volume		0.000	03:30	
Air Blank		0.000	03:31	
Air Blank		0.000	03:33	
Subject Sample	# 2	0.254	03:33	
Breath Volume		1.183	03:33	
Air Blank		0.000	03:34	
Air Blank		0.000	03:36	
Subject Sample	# 3	0.245	03:39	
Breath Volume		2.171	03:39	
Air Blank		0.000	03:40	
Control Test		0.079	03:40	
Air Blank		0.000	03:40	
Diagnostics Check		OK	03:41	

Data Download Date/Time: 11/13/2019 10:58

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MAZZELLA DI BOSCO,DOMENICO
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 10:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:49

SUBJECT'S LAST NAME: THORNTON

DATE/TIME OF TEST: 11/03/2019 13:22:41

Diagnostics Check		OK	13:16
Air Blank		0.000	13:16
Control Test		0.079	13:17
Air Blank		0.000	13:17
Subject Sample	# 1	0.000	13:18
Breath Volume		2.105	13:18
Air Blank		0.000	13:18
Air Blank		0.000	13:20
Subject Sample	# 2	0.000	13:21
Breath Volume		1.253	13:21
Air Blank		0.000	13:21
Control Test		0.079	13:22
Air Blank		0.000	13:22
Diagnostics Check		OK	13:22

Data Download Date/Time: 11/13/2019 10:58

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KELLIHER,MICHAEL
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 18:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:34

SUBJECT'S LAST NAME: PACHECO

DATE/TIME OF TEST: 11/03/2019 20:12:07

Diagnostics Check		OK	19:57	
Air Blank		0.000	19:58	
Control Test		0.080	19:58	
Air Blank		0.000	19:58	
Subject Sample	# 1	0.169	20:00	
Breath Volume		1.718	20:00	
Air Blank		0.000	20:01	
Air Blank		0.000	20:03	
Subject Sample	# 2	0.131	20:06	Volume Not Met
Breath Volume		0.218	20:06	
Air Blank		0.000	20:07	
Air Blank		0.000	20:08	
Subject Sample	# 3	0.172	20:10	
Breath Volume		2.046	20:10	
Air Blank		0.000	20:11	
Control Test		0.077	20:11	
Air Blank		0.000	20:12	
Diagnostics Check		OK	20:12	

Data Download Date/Time: 11/13/2019 10:58

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KELLIHER,MICHAEL
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: LESPINASSE

DATE/TIME OF TEST: 11/04/2019 01:28:54

Diagnostics Check		OK	01:22
Air Blank		0.000	01:23
Control Test		0.079	01:23
Air Blank		0.000	01:23
Subject Sample	# 1	0.154	01:24
Breath Volume		1.789	01:24
Air Blank		0.000	01:25
Air Blank		0.000	01:26
Subject Sample	# 2	0.151	01:27
Breath Volume		1.582	01:27
Air Blank		0.000	01:27
Control Test		0.079	01:28
Air Blank		0.000	01:28
Diagnostics Check		OK	01:28

Data Download Date/Time: 11/13/2019 10:58

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ESCOBAR,DIEGO
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 14:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:10

SUBJECT'S LAST NAME: VASQUEZ

DATE/TIME OF TEST: 11/04/2019 15:48:37

Diagnostics Check		OK	15:42
Air Blank		0.000	15:42
Control Test		0.080	15:43
Air Blank		0.000	15:43
Subject Sample	# 1	0.082	15:44
Breath Volume		3.261	15:44
Air Blank		0.000	15:44
Air Blank		0.000	15:46
Subject Sample	# 2	0.087	15:46
Breath Volume		3.105	15:46
Air Blank		0.000	15:47
Control Test		0.080	15:48
Air Blank		0.000	15:48
Diagnostics Check		OK	15:48

Data Download Date/Time: 11/13/2019 10:58

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SCHMERTZLER,KEVIN
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: MOUNTS

DATE/TIME OF TEST: 11/05/2019 22:22:47

Diagnostics Check		OK	22:11	
Air Blank		0.000	22:11	
Control Test		0.080	22:12	
Air Blank		0.000	22:12	
Subject Sample	# 1		22:15	No Sample Provided
Breath Volume		0.000	22:15	
Air Blank		0.000	22:16	
Air Blank		0.000	22:18	
Subject Sample	# 2		22:21	No Sample Provided
Breath Volume		0.000	22:21	
Air Blank		0.000	22:21	
Control Test		0.080	22:22	
Air Blank		0.000	22:22	
Diagnostics Check		OK	22:22	

Data Download Date/Time: 11/13/2019 10:58

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: DUPERVAL

DATE/TIME OF TEST: 11/08/2019 03:40:51

Diagnostics Check		OK	03:34
Air Blank		0.000	03:34
Control Test		0.080	03:34
Air Blank		0.000	03:35
Subject Sample	# 1	0.210	03:36
Breath Volume		2.269	03:36
Air Blank		0.000	03:36
Air Blank		0.000	03:38
Subject Sample	# 2	0.216	03:39
Breath Volume		2.300	03:39
Air Blank		0.000	03:39
Control Test		0.078	03:40
Air Blank		0.000	03:40
Diagnostics Check		OK	03:40

Data Download Date/Time: 11/13/2019 10:58

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: YEARWOOD,VANINA
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 06:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:30

SUBJECT'S LAST NAME: BOLANOS

DATE/TIME OF TEST: 11/08/2019 08:12:38

Diagnostics Check		OK	08:02
Air Blank		0.000	08:03
Control Test		0.079	08:03
Air Blank		0.000	08:04
Subject Sample	# 1	0.292	08:04
Breath Volume		2.671	08:04
Air Blank		0.000	08:05
Air Blank		0.000	08:07
Subject Sample	# 2	0.265	08:07
Breath Volume		1.378	08:07
Air Blank		0.000	08:08
Air Blank		0.000	08:10
Subject Sample	# 3	0.301	08:10
Breath Volume		1.894	08:10
Air Blank		0.000	08:11
Control Test		0.079	08:12
Air Blank		0.000	08:12
Diagnostics Check		OK	08:12

Data Download Date/Time: 11/13/2019 10:58

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAPP,JOSHUA,R
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: TOLEDO,S
ARRESTING OFFICER AGENCY: Hallandale Beach PD
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 12:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:45

SUBJECT'S LAST NAME: THERRELL

DATE/TIME OF TEST: 11/10/2019 17:32:30

Diagnostics Check		OK	17:25
Air Blank		0.000	17:26
Control Test		0.080	17:26
Air Blank		0.000	17:27
Subject Sample	# 1	0.161	17:28
Breath Volume		2.300	17:28
Air Blank		0.000	17:28
Air Blank		0.000	17:30
Subject Sample	# 2	0.160	17:30
Breath Volume		1.167	17:30
Air Blank		0.000	17:31
Control Test		0.080	17:31
Air Blank		0.000	17:32
Diagnostics Check		OK	17:32

Data Download Date/Time: 11/13/2019 10:58

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 19:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: HUNTT

DATE/TIME OF TEST: 11/12/2019 20:55:19

Diagnostics Check		OK	20:48
Air Blank		0.000	20:48
Control Test		0.080	20:49
Air Blank		0.000	20:49
Subject Sample	# 1	0.130	20:50
Breath Volume		2.136	20:50
Air Blank		0.000	20:50
Air Blank		0.000	20:52
Subject Sample	# 2	0.132	20:53
Breath Volume		2.535	20:53
Air Blank		0.000	20:54
Control Test		0.080	20:54
Air Blank		0.000	20:55
Diagnostics Check		OK	20:55

Data Download Date/Time: 11/13/2019 10:58

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 19:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: OSSELIN

DATE/TIME OF TEST: 11/13/2019 20:42:02

Diagnostics Check		OK	20:30	
Air Blank		0.000	20:30	
Control Test		0.080	20:30	
Air Blank		0.000	20:31	
Subject Sample	# 1	0.153	20:34	Volume Not Met
Breath Volume		1.015	20:34	
Air Blank		0.000	20:35	
Air Blank		0.000	20:37	
Subject Sample	# 2	0.146	20:40	Volume Not Met
Breath Volume		0.429	20:40	
Air Blank		0.000	20:41	
Control Test		0.077	20:41	
Air Blank		0.000	20:41	
Diagnostics Check		OK	20:42	

Data Download Date/Time: 12/05/2019 12:08

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DERIET,RAFAEL
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: BITTER

DATE/TIME OF TEST: 11/14/2019 01:37:50

Diagnostics Check		OK	01:34	
Air Blank		0.000	01:34	
Control Test		0.079	01:35	
Air Blank		0.000	01:35	
Subject Sample	# 1		01:36	Interferent Detect
Breath Volume		2.578	01:36	
Air Blank		0.000	01:36	
Control Test		0.078	01:37	
Air Blank		0.000	01:37	
Diagnostics Check		OK	01:37	

Data Download Date/Time: 12/05/2019 12:08

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DERIET,RAFAEL
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: BITTER

DATE/TIME OF TEST: 11/14/2019 01:43:13

Diagnostics Check		OK	01:39	
Air Blank		0.000	01:40	
Control Test		0.079	01:40	
Air Blank		0.000	01:41	
Subject Sample	# 1		01:41	Interferent Detect
Breath Volume		2.261	01:41	
Air Blank		0.000	01:42	
Control Test		0.078	01:42	
Air Blank		0.000	01:43	
Diagnostics Check		OK	01:43	

Data Download Date/Time: 12/05/2019 12:08

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,YVETTE,D
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FOWLER,SHAUN
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 13:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:15

SUBJECT'S LAST NAME: DOVO

DATE/TIME OF TEST: 11/14/2019 14:49:55

Diagnostics Check		OK	14:41
Air Blank		0.000	14:42
Control Test		0.080	14:42
Air Blank		0.000	14:43
Subject Sample	# 1	0.000	14:43
Breath Volume		1.187	14:43
Air Blank		0.000	14:44
Air Blank		0.000	14:46
Subject Sample	# 2	0.000	14:48
Breath Volume		1.359	14:48
Air Blank		0.000	14:48
Control Test		0.081	14:49
Air Blank		0.000	14:49
Diagnostics Check		OK	14:49

Data Download Date/Time: 12/05/2019 12:08

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAPP,JOSHUA,R
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ODOR,R
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 04:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:56

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 11/15/2019 05:24:28

Diagnostics Check		OK	05:16
Air Blank		0.000	05:17
Control Test		0.079	05:17
Air Blank		0.000	05:17
Subject Sample	# 1	0.066	05:19
Breath Volume		1.765	05:19
Air Blank		0.000	05:19
Air Blank		0.000	05:21
Subject Sample	# 2	0.061	05:22
Breath Volume		1.621	05:22
Air Blank		0.000	05:23
Control Test		0.079	05:23
Air Blank		0.000	05:24
Diagnostics Check		OK	05:24

Data Download Date/Time: 12/05/2019 12:08

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAPP,JOSHUA,R
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HOLLOWAY,D
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 04:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:45

SUBJECT'S LAST NAME: BOUTIN

DATE/TIME OF TEST: 11/15/2019 06:12:11

Diagnostics Check		OK	06:05
Air Blank		0.000	06:06
Control Test		0.080	06:06
Air Blank		0.000	06:07
Subject Sample	# 1	0.084	06:07
Breath Volume		1.625	06:07
Air Blank		0.000	06:08
Air Blank		0.000	06:10
Subject Sample	# 2	0.090	06:10
Breath Volume		2.261	06:10
Air Blank		0.000	06:11
Control Test		0.079	06:11
Air Blank		0.000	06:12
Diagnostics Check		OK	06:12

Data Download Date/Time: 12/05/2019 12:08

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LUNA,KRISTINA
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: PALMER

DATE/TIME OF TEST: 11/15/2019 23:43:20

Diagnostics Check		OK	23:36
Air Blank		0.000	23:37
Control Test		0.079	23:37
Air Blank		0.000	23:37
Subject Sample	# 1	0.126	23:38
Breath Volume		2.558	23:38
Air Blank		0.000	23:39
Air Blank		0.000	23:41
Subject Sample	# 2	0.125	23:41
Breath Volume		1.605	23:41
Air Blank		0.000	23:42
Control Test		0.079	23:42
Air Blank		0.000	23:43
Diagnostics Check		OK	23:43

Data Download Date/Time: 12/05/2019 12:08

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: O'DOR,ROBERT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: BOCCADISI

DATE/TIME OF TEST: 11/16/2019 03:26:38

Diagnostics Check		OK	03:20
Air Blank		0.000	03:20
Control Test		0.080	03:21
Air Blank		0.000	03:21
Subject Sample	# 1	0.221	03:22
Breath Volume		1.925	03:22
Air Blank		0.000	03:22
Air Blank		0.000	03:24
Subject Sample	# 2	0.235	03:24
Breath Volume		2.792	03:24
Air Blank		0.000	03:25
Control Test		0.080	03:26
Air Blank		0.000	03:26
Diagnostics Check		OK	03:26

Data Download Date/Time: 12/05/2019 12:08

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: SHARP

DATE/TIME OF TEST: 11/16/2019 03:51:27

Diagnostics Check		OK	03:44
Air Blank		0.000	03:44
Control Test		0.079	03:45
Air Blank		0.000	03:45
Subject Sample	# 1	0.129	03:46
Breath Volume		1.695	03:46
Air Blank		0.000	03:46
Air Blank		0.000	03:48
Subject Sample	# 2	0.136	03:49
Breath Volume		2.914	03:49
Air Blank		0.000	03:50
Control Test		0.079	03:50
Air Blank		0.000	03:51
Diagnostics Check		OK	03:51

Data Download Date/Time: 12/05/2019 12:08

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: STONER,NATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 04:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:13

SUBJECT'S LAST NAME: TRAZILIEN

DATE/TIME OF TEST: 11/16/2019 05:45:37

Diagnostics Check		OK	05:37
Air Blank		0.000	05:38
Control Test		0.080	05:38
Air Blank		0.000	05:38
Subject Sample	# 1	0.173	05:40
Breath Volume		1.949	05:40
Air Blank		0.000	05:41
Air Blank		0.000	05:43
Subject Sample	# 2	0.173	05:43
Breath Volume		1.761	05:43
Air Blank		0.000	05:44
Control Test		0.077	05:45
Air Blank		0.000	05:45
Diagnostics Check		OK	05:45

Data Download Date/Time: 12/05/2019 12:08

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 22:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:01

SUBJECT'S LAST NAME: GREGORY

DATE/TIME OF TEST: 11/16/2019 23:37:42

Diagnostics Check		OK	23:24	
Air Blank		0.000	23:25	
Control Test		0.079	23:25	
Air Blank		0.000	23:26	
Subject Sample	# 1	0.050	23:29	Volume Not Met
Breath Volume		0.542	23:29	
Air Blank		0.000	23:30	
Air Blank		0.000	23:31	
Subject Sample	# 2	0.068	23:32	
Breath Volume		1.597	23:32	
Air Blank		0.000	23:33	
Air Blank		0.000	23:35	
Subject Sample	# 3	0.069	23:36	
Breath Volume		1.574	23:36	
Air Blank		0.000	23:36	
Control Test		0.079	23:37	
Air Blank		0.000	23:37	
Diagnostics Check		OK	23:37	

Data Download Date/Time: 12/05/2019 12:08

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BOONE,MICHAEL
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 04:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:28

SUBJECT'S LAST NAME: PEREZ JR

DATE/TIME OF TEST: 11/17/2019 06:12:30

Diagnostics Check		OK	06:05
Air Blank		0.000	06:06
Control Test		0.080	06:06
Air Blank		0.000	06:07
Subject Sample	# 1	0.156	06:07
Breath Volume		2.757	06:07
Air Blank		0.000	06:08
Air Blank		0.000	06:10
Subject Sample	# 2	0.161	06:10
Breath Volume		3.062	06:10
Air Blank		0.000	06:11
Control Test		0.078	06:11
Air Blank		0.000	06:12
Diagnostics Check		OK	06:12

Data Download Date/Time: 12/05/2019 12:08

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FRESNILLO,JULIETTE
ARRESTING OFFICER AGENCY: Hallandale Beach PD
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 05:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:53

SUBJECT'S LAST NAME: VECCHIO

DATE/TIME OF TEST: 11/17/2019 06:36:08

Diagnostics Check		OK	06:29
Air Blank		0.000	06:30
Control Test		0.078	06:30
Air Blank		0.000	06:31
Subject Sample	# 1	0.000	06:31
Breath Volume		1.625	06:31
Air Blank		0.000	06:32
Air Blank		0.000	06:34
Subject Sample	# 2	0.000	06:34
Breath Volume		2.242	06:34
Air Blank		0.000	06:35
Control Test		0.080	06:35
Air Blank		0.000	06:36
Diagnostics Check		OK	06:36

Data Download Date/Time: 12/05/2019 12:08

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ROJAS,RANDY
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 05:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:06

SUBJECT'S LAST NAME: TORRES SUAREZ

DATE/TIME OF TEST: 11/17/2019 06:48:16

Diagnostics Check		OK	06:41
Air Blank		0.000	06:42
Control Test		0.079	06:42
Air Blank		0.000	06:43
Subject Sample	# 1	0.183	06:43
Breath Volume		3.214	06:43
Air Blank		0.000	06:44
Air Blank		0.000	06:45
Subject Sample	# 2	0.167	06:46
Breath Volume		1.585	06:46
Air Blank		0.000	06:47
Control Test		0.079	06:47
Air Blank		0.000	06:48
Diagnostics Check		OK	06:48

Data Download Date/Time: 12/05/2019 12:08

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: VARGAS,JOSE
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 10:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:32

SUBJECT'S LAST NAME: RAMIREZ BRITO

DATE/TIME OF TEST: 11/17/2019 12:59:19

Diagnostics Check		OK	12:52
Air Blank		0.000	12:52
Control Test		0.080	12:53
Air Blank		0.000	12:53
Subject Sample	# 1	0.157	12:54
Breath Volume		1.847	12:54
Air Blank		0.000	12:55
Air Blank		0.000	12:57
Subject Sample	# 2	0.162	12:57
Breath Volume		2.093	12:57
Air Blank		0.000	12:58
Control Test		0.079	12:58
Air Blank		0.000	12:59
Diagnostics Check		OK	12:59

Data Download Date/Time: 12/05/2019 12:08

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KELLIHER,MICHAEL
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 21:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: PIERRE

DATE/TIME OF TEST: 11/17/2019 22:04:28

Diagnostics Check		OK	22:00	
Air Blank		0.000	22:01	
Control Test		0.080	22:01	
Air Blank		0.000	22:02	
Subject Sample	# 1		22:02	Slope Not Met
Breath Volume		1.011	22:02	
Air Blank		0.000	22:03	
Control Test		0.077	22:03	
Air Blank		0.000	22:04	
Diagnostics Check		OK	22:04	

Data Download Date/Time: 12/05/2019 12:08

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KELLIHER,MICHAEL
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 21:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: PIERRE

DATE/TIME OF TEST: 11/17/2019 22:32:46

Diagnostics Check		OK	22:26
Air Blank		0.000	22:27
Control Test		0.079	22:27
Air Blank		0.000	22:27
Subject Sample	# 1	0.125	22:28
Breath Volume		1.714	22:28
Air Blank		0.000	22:29
Air Blank		0.000	22:30
Subject Sample	# 2	0.123	22:31
Breath Volume		1.492	22:31
Air Blank		0.000	22:31
Control Test		0.079	22:32
Air Blank		0.000	22:32
Diagnostics Check		OK	22:32

Data Download Date/Time: 12/05/2019 12:08

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: STONER,NATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: CROPPER

DATE/TIME OF TEST: 11/20/2019 01:26:38

Diagnostics Check		OK	01:17
Air Blank		0.000	01:17
Control Test		0.080	01:18
Air Blank		0.000	01:18
Subject Sample	# 1	0.140	01:21
Breath Volume		2.523	01:21
Air Blank		0.000	01:22
Air Blank		0.000	01:23
Subject Sample	# 2	0.124	01:24
Breath Volume		1.156	01:24
Air Blank		0.000	01:25
Control Test		0.078	01:26
Air Blank		0.000	01:26
Diagnostics Check		OK	01:26

Data Download Date/Time: 12/05/2019 12:08

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: STONER,NATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: BLAKEY

DATE/TIME OF TEST: 11/22/2019 04:46:42

Diagnostics Check		OK	04:39
Air Blank		0.000	04:39
Control Test		0.080	04:40
Air Blank		0.000	04:40
Subject Sample	# 1	0.110	04:41
Breath Volume		1.867	04:41
Air Blank		0.000	04:42
Air Blank		0.000	04:44
Subject Sample	# 2	0.112	04:45
Breath Volume		1.148	04:45
Air Blank		0.000	04:45
Control Test		0.079	04:46
Air Blank		0.000	04:46
Diagnostics Check		OK	04:46

Data Download Date/Time: 12/05/2019 12:08

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ROJAS,RANDY
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 04:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: LALLANCE

DATE/TIME OF TEST: 11/22/2019 05:04:57

Diagnostics Check		OK	04:58
Air Blank		0.000	04:58
Control Test		0.079	04:59
Air Blank		0.000	04:59
Subject Sample	# 1	0.113	05:00
Breath Volume		1.238	05:00
Air Blank		0.000	05:00
Air Blank		0.000	05:02
Subject Sample	# 2	0.116	05:03
Breath Volume		1.632	05:03
Air Blank		0.000	05:04
Control Test		0.080	05:04
Air Blank		0.000	05:04
Diagnostics Check		OK	05:04

Data Download Date/Time: 12/05/2019 12:08

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LUNA,KRISTINA
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: VIRGIN

DATE/TIME OF TEST: 11/23/2019 00:16:08

Diagnostics Check		OK	00:08
Air Blank		0.000	00:09
Control Test		0.080	00:09
Air Blank		0.000	00:10
Subject Sample	# 1	0.150	00:10
Breath Volume		2.074	00:10
Air Blank		0.000	00:11
Air Blank		0.000	00:13
Subject Sample	# 2	0.153	00:14
Breath Volume		2.496	00:14
Air Blank		0.000	00:15
Control Test		0.079	00:15
Air Blank		0.000	00:16
Diagnostics Check		OK	00:16

Data Download Date/Time: 12/05/2019 12:08

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: ZENZ

DATE/TIME OF TEST: 11/23/2019 03:58:08

Diagnostics Check		OK	03:46	
Air Blank		0.000	03:47	
Control Test		0.080	03:47	
Air Blank		0.000	03:48	
Subject Sample	# 1		03:51	No Sample Provided
Breath Volume		0.000	03:51	
Air Blank		0.000	03:51	
Air Blank		0.000	03:53	
Subject Sample	# 2		03:56	No Sample Provided
Breath Volume		0.000	03:56	
Air Blank		0.000	03:57	
Control Test		0.080	03:57	
Air Blank		0.000	03:58	
Diagnostics Check		OK	03:58	

Data Download Date/Time: 12/05/2019 12:08

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DEOLIVEIRA,JULIANA
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 11:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:03

SUBJECT'S LAST NAME: LICENA

DATE/TIME OF TEST: 11/23/2019 13:42:37

Diagnostics Check		OK	13:33
Air Blank		0.000	13:34
Control Test		0.080	13:34
Air Blank		0.000	13:35
Subject Sample	# 1	0.249	13:37
Breath Volume		1.390	13:37
Air Blank		0.000	13:38
Air Blank		0.000	13:40
Subject Sample	# 2	0.246	13:40
Breath Volume		1.593	13:40
Air Blank		0.000	13:41
Control Test		0.079	13:41
Air Blank		0.000	13:42
Diagnostics Check		OK	13:42

Data Download Date/Time: 12/05/2019 12:08

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HOLLOWAY,DEVARIOUS
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: GUEVARA

DATE/TIME OF TEST: 11/24/2019 03:10:15

Diagnostics Check		OK	03:03
Air Blank		0.000	03:04
Control Test		0.080	03:04
Air Blank		0.000	03:05
Subject Sample	# 1	0.088	03:05
Breath Volume		2.027	03:05
Air Blank		0.000	03:06
Air Blank		0.000	03:08
Subject Sample	# 2	0.089	03:08
Breath Volume		1.707	03:08
Air Blank		0.000	03:09
Control Test		0.079	03:09
Air Blank		0.000	03:10
Diagnostics Check		OK	03:10

Data Download Date/Time: 12/05/2019 12:08

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KAPLAN,SAM
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: SAMUEL

DATE/TIME OF TEST: 11/24/2019 04:10:55

Diagnostics Check		OK	04:04
Air Blank		0.000	04:04
Control Test		0.079	04:04
Air Blank		0.000	04:05
Subject Sample	# 1	0.277	04:06
Breath Volume		1.437	04:06
Air Blank		0.000	04:06
Air Blank		0.000	04:08
Subject Sample	# 2	0.265	04:09
Breath Volume		1.562	04:09
Air Blank		0.000	04:09
Control Test		0.078	04:10
Air Blank		0.000	04:10
Diagnostics Check		OK	04:10

Data Download Date/Time: 12/05/2019 12:08

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: WILSON,LAMAR
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 20:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:38

SUBJECT'S LAST NAME: CLAUSEN

DATE/TIME OF TEST: 11/25/2019 21:07:47

Diagnostics Check		OK	21:01
Air Blank		0.000	21:01
Control Test		0.079	21:01
Air Blank		0.000	21:02
Subject Sample	# 1	0.179	21:02
Breath Volume		2.117	21:02
Air Blank		0.000	21:03
Air Blank		0.000	21:05
Subject Sample	# 2	0.177	21:06
Breath Volume		2.878	21:06
Air Blank		0.000	21:06
Control Test		0.079	21:07
Air Blank		0.000	21:07
Diagnostics Check		OK	21:07

Data Download Date/Time: 12/05/2019 12:08

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HORST,S.R.
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: GOFF

DATE/TIME OF TEST: 11/26/2019 01:42:59

Diagnostics Check		OK	01:35
Air Blank		0.000	01:35
Control Test		0.080	01:35
Air Blank		0.000	01:36
Subject Sample	# 1	0.158	01:37
Breath Volume		2.511	01:37
Air Blank		0.000	01:38
Air Blank		0.000	01:40
Subject Sample	# 2	0.152	01:41
Breath Volume		2.171	01:41
Air Blank		0.000	01:42
Control Test		0.078	01:42
Air Blank		0.000	01:42
Diagnostics Check		OK	01:42

Data Download Date/Time: 12/05/2019 12:08

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: PACK,JEFFREY
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 11/26/2019 03:50:00

Diagnostics Check		OK	03:43
Air Blank		0.000	03:43
Control Test		0.079	03:44
Air Blank		0.000	03:44
Subject Sample	# 1	0.175	03:45
Breath Volume		3.761	03:45
Air Blank		0.000	03:45
Air Blank		0.000	03:47
Subject Sample	# 2	0.181	03:48
Breath Volume		3.238	03:48
Air Blank		0.000	03:49
Control Test		0.079	03:49
Air Blank		0.000	03:49
Diagnostics Check		OK	03:50

Data Download Date/Time: 12/05/2019 12:08

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: DUCHESNE

DATE/TIME OF TEST: 11/26/2019 23:25:00

Diagnostics Check		OK	23:18
Air Blank		0.000	23:18
Control Test		0.080	23:18
Air Blank		0.000	23:19
Subject Sample	# 1	0.154	23:20
Breath Volume		3.109	23:20
Air Blank		0.000	23:20
Air Blank		0.000	23:22
Subject Sample	# 2	0.149	23:23
Breath Volume		3.582	23:23
Air Blank		0.000	23:24
Control Test		0.079	23:24
Air Blank		0.000	23:24
Diagnostics Check		OK	23:25

Data Download Date/Time: 12/05/2019 12:08

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: JIMENEZ,JOSE
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 17:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:53

SUBJECT'S LAST NAME: CESPEDES

DATE/TIME OF TEST: 11/27/2019 19:27:49

Diagnostics Check		OK	19:14	
Air Blank		0.000	19:14	
Control Test		0.080	19:14	
Air Blank		0.000	19:15	
Subject Sample	# 1		19:18	No Sample Provided
Breath Volume		0.000	19:18	
Air Blank		0.000	19:18	
Air Blank		0.000	19:21	
Subject Sample	# 2	0.198	19:22	
Breath Volume		1.605	19:22	
Air Blank		0.000	19:23	
Air Blank		0.000	19:25	
Subject Sample	# 3	0.212	19:26	
Breath Volume		2.453	19:26	
Air Blank		0.000	19:26	
Control Test		0.080	19:27	
Air Blank		0.000	19:27	
Diagnostics Check		OK	19:27	

Data Download Date/Time: 12/05/2019 12:08

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BOHM,JONATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 20:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:33

SUBJECT'S LAST NAME: KUCHARSKI

DATE/TIME OF TEST: 11/28/2019 22:03:23

Diagnostics Check		OK	21:53
Air Blank		0.000	21:54
Control Test		0.080	21:54
Air Blank		0.000	21:55
Subject Sample	# 1	0.199	21:55
Breath Volume		2.207	21:55
Air Blank		0.000	21:56
Air Blank		0.000	21:58
Subject Sample	# 2	0.220	21:58
Breath Volume		2.316	21:58
Air Blank		0.000	21:59
Air Blank		0.000	22:01
Subject Sample	# 3	0.206	22:01
Breath Volume		2.328	22:01
Air Blank		0.000	22:02
Control Test		0.080	22:02
Air Blank		0.000	22:03
Diagnostics Check		OK	22:03

Data Download Date/Time: 12/05/2019 12:08

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: STONER,NATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: HEAVEN

DATE/TIME OF TEST: 11/29/2019 00:40:33

Diagnostics Check		OK	00:34
Air Blank		0.000	00:34
Control Test		0.080	00:34
Air Blank		0.000	00:35
Subject Sample	# 1	0.128	00:35
Breath Volume		1.601	00:35
Air Blank		0.000	00:36
Air Blank		0.000	00:38
Subject Sample	# 2	0.146	00:38
Breath Volume		1.515	00:38
Air Blank		0.000	00:39
Control Test		0.079	00:39
Air Blank		0.000	00:40
Diagnostics Check		OK	00:40

Data Download Date/Time: 12/05/2019 12:09

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: NEWELL

DATE/TIME OF TEST: 11/29/2019 03:40:38

Diagnostics Check		OK	03:33
Air Blank		0.000	03:33
Control Test		0.080	03:34
Air Blank		0.000	03:34
Subject Sample	# 1	0.227	03:35
Breath Volume		1.332	03:35
Air Blank		0.000	03:36
Air Blank		0.000	03:38
Subject Sample	# 2	0.213	03:38
Breath Volume		1.613	03:38
Air Blank		0.000	03:39
Control Test		0.079	03:40
Air Blank		0.000	03:40
Diagnostics Check		OK	03:40

Data Download Date/Time: 12/05/2019 12:09

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAPP,JOSHUA,R
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: WILSON,L
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 16:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:13

SUBJECT'S LAST NAME: KEYTE

DATE/TIME OF TEST: 11/30/2019 17:56:00

Diagnostics Check		OK	17:47
Air Blank		0.000	17:47
Control Test		0.080	17:48
Air Blank		0.000	17:48
Subject Sample	# 1	0.000	17:51
Breath Volume		2.390	17:51
Air Blank		0.000	17:51
Air Blank		0.000	17:53
Subject Sample	# 2	0.000	17:54
Breath Volume		1.238	17:54
Air Blank		0.000	17:55
Control Test		0.081	17:55
Air Blank		0.000	17:55
Diagnostics Check		OK	17:56

Data Download Date/Time: 12/05/2019 12:09

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FITHIAN,JASON,
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: FITHIAN,JASON
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 11/05/2019 05:32:17

Diagnostics Check		OK	05:25
Air Blank		0.000	05:25
Control Test		0.077	05:26
Air Blank		0.000	05:26
Subject Sample	# 1	0.000	05:27
Breath Volume		2.312	05:27
Air Blank		0.000	05:28
Air Blank		0.000	05:30
Subject Sample	# 2	0.000	05:30
Breath Volume		2.078	05:30
Air Blank		0.000	05:31
Control Test		0.076	05:31
Air Blank		0.000	05:32
Diagnostics Check		OK	05:32

Data Download Date/Time: 11/13/2019 12:28

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FITHIAN,JAY,
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: KEATING,KAYLA
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:37

SUBJECT'S LAST NAME: GRANDINETTI

DATE/TIME OF TEST: 11/17/2019 22:15:05

Diagnostics Check		OK	22:08
Air Blank		0.000	22:08
Control Test		0.078	22:08
Air Blank		0.000	22:09
Subject Sample	# 1	0.173	22:10
Breath Volume		1.699	22:10
Air Blank		0.000	22:10
Air Blank		0.000	22:12
Subject Sample	# 2	0.174	22:13
Breath Volume		2.371	22:13
Air Blank		0.000	22:14
Control Test		0.079	22:14
Air Blank		0.000	22:15
Diagnostics Check		OK	22:15

Data Download Date/Time: 12/05/2019 12:56

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FITHIAN,JAY,
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: KEATING,KAYLA
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:37

SUBJECT'S LAST NAME: WHITED

DATE/TIME OF TEST: 11/24/2019 20:04:28

Diagnostics Check		OK	19:58
Air Blank		0.000	19:58
Control Test		0.078	19:58
Air Blank		0.000	19:59
Subject Sample	# 1	0.231	19:59
Breath Volume		3.750	19:59
Air Blank		0.000	20:00
Air Blank		0.000	20:02
Subject Sample	# 2	0.230	20:02
Breath Volume		4.480	20:02
Air Blank		0.000	20:03
Control Test		0.078	20:03
Air Blank		0.000	20:04
Diagnostics Check		OK	20:04

Data Download Date/Time: 12/05/2019 12:56

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FITHIAN,JASON,
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: FITHIAN,J
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: CRAWLEY

DATE/TIME OF TEST: 11/26/2019 00:48:30

Diagnostics Check		OK	00:41
Air Blank		0.000	00:42
Control Test		0.077	00:42
Air Blank		0.000	00:42
Subject Sample	# 1	0.000	00:43
Breath Volume		2.054	00:43
Air Blank		0.000	00:44
Air Blank		0.000	00:46
Subject Sample	# 2	0.000	00:47
Breath Volume		2.195	00:47
Air Blank		0.000	00:47
Control Test		0.078	00:47
Air Blank		0.000	00:48
Diagnostics Check		OK	00:48

Data Download Date/Time: 12/05/2019 12:56

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FITHIAN,JAY,
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: KEATING,KAYLA
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: PAYAN-SERRATOS

DATE/TIME OF TEST: 11/28/2019 00:29:29

Diagnostics Check		OK	00:22
Air Blank		0.000	00:22
Control Test		0.077	00:23
Air Blank		0.000	00:23
Subject Sample	# 1	0.115	00:24
Breath Volume		1.574	00:24
Air Blank		0.000	00:25
Air Blank		0.000	00:26
Subject Sample	# 2	0.132	00:27
Breath Volume		2.355	00:27
Air Blank		0.000	00:28
Control Test		0.079	00:28
Air Blank		0.000	00:29
Diagnostics Check		OK	00:29

Data Download Date/Time: 12/05/2019 12:56

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAGE,L,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: STAUFFER,.
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 11/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 15:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:30

SUBJECT'S LAST NAME: REFFRUSCHINNI

DATE/TIME OF TEST: 11/01/2019 15:59:39

Diagnostics Check		OK	15:53
Air Blank		0.000	15:53
Control Test		0.081	15:54
Air Blank		0.000	15:54
Subject Sample	# 1	0.089	15:55
Breath Volume		1.476	15:55
Air Blank		0.000	15:55
Air Blank		0.000	15:57
Subject Sample	# 2	0.091	15:58
Breath Volume		1.351	15:58
Air Blank		0.000	15:58
Control Test		0.080	15:59
Air Blank		0.000	15:59
Diagnostics Check		OK	15:59

Data Download Date/Time: 12/03/2019 16:40

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOODMAN,R.,R
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: COMBS,S.
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 11/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 12:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:37

SUBJECT'S LAST NAME: GALLUS

DATE/TIME OF TEST: 11/02/2019 14:08:14

Diagnostics Check		OK	14:01
Air Blank		0.000	14:02
Control Test		0.082	14:02
Air Blank		0.000	14:03
Subject Sample	# 1	0.030	14:03
Breath Volume		2.207	14:03
Air Blank		0.000	14:04
Air Blank		0.000	14:06
Subject Sample	# 2	0.029	14:06
Breath Volume		1.894	14:06
Air Blank		0.000	14:07
Control Test		0.081	14:07
Air Blank		0.000	14:08
Diagnostics Check		OK	14:08

Data Download Date/Time: 12/03/2019 16:40

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUONPANE,V,D
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: SOLIS,E
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 11/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: BELINA

DATE/TIME OF TEST: 11/03/2019 02:55:24

Diagnostics Check		OK	02:45	
Air Blank		0.000	02:46	
Control Test		0.082	02:46	
Air Blank		0.000	02:47	
Subject Sample	# 1	0.129	02:49	
Breath Volume		2.613	02:49	
Air Blank		0.000	02:50	
Air Blank		0.000	02:51	
Subject Sample	# 2		02:53	Subject Test Refused
Breath Volume		0.097	02:53	
Air Blank		0.000	02:54	
Control Test		0.079	02:54	
Air Blank		0.000	02:55	
Diagnostics Check		OK	02:55	

Data Download Date/Time: 12/03/2019 16:40

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOPS,C,M
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: BENOIT,J
ARRESTING OFFICER AGENCY: FHP TROOP F
LAST AGENCY INSPECTION: 11/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: MEHRA

DATE/TIME OF TEST: 11/04/2019 19:23:00

Diagnostics Check		OK	19:16
Air Blank		0.000	19:17
Control Test		0.082	19:17
Air Blank		0.000	19:18
Subject Sample	# 1	0.000	19:18
Breath Volume		2.238	19:18
Air Blank		0.000	19:19
Air Blank		0.000	19:21
Subject Sample	# 2	0.000	19:21
Breath Volume		1.503	19:21
Air Blank		0.000	19:22
Control Test		0.081	19:22
Air Blank		0.000	19:22
Diagnostics Check		OK	19:23

Data Download Date/Time: 12/03/2019 16:40

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODI,DOMENIC,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: WALBERT,A
ARRESTING OFFICER AGENCY: NAPLES PD
LAST AGENCY INSPECTION: 11/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: FULLER

DATE/TIME OF TEST: 11/05/2019 05:14:32

Diagnostics Check		OK	05:08
Air Blank		0.000	05:08
Control Test		0.081	05:08
Air Blank		0.000	05:09
Subject Sample	# 1	0.271	05:09
Breath Volume		1.984	05:09
Air Blank		0.000	05:10
Air Blank		0.000	05:12
Subject Sample	# 2	0.270	05:12
Breath Volume		2.472	05:12
Air Blank		0.000	05:13
Control Test		0.081	05:13
Air Blank		0.000	05:14
Diagnostics Check		OK	05:14

Data Download Date/Time: 12/03/2019 16:40

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TARR,L,J
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: SUEROLANTIGUA,J
ARRESTING OFFICER AGENCY: FHP TROOP F
LAST AGENCY INSPECTION: 11/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: GONZALES

DATE/TIME OF TEST: 11/06/2019 01:29:31

Diagnostics Check		OK	01:22
Air Blank		0.000	01:23
Control Test		0.082	01:23
Air Blank		0.000	01:24
Subject Sample	# 1	0.000	01:25
Breath Volume		2.671	01:25
Air Blank		0.000	01:25
Air Blank		0.000	01:27
Subject Sample	# 2	0.000	01:28
Breath Volume		3.328	01:28
Air Blank		0.000	01:28
Control Test		0.081	01:28
Air Blank		0.000	01:29
Diagnostics Check		OK	01:29

Data Download Date/Time: 12/03/2019 16:40

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOODMAN,R.,R
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: BIRTOLO,J.
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 11/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 11:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:06

SUBJECT'S LAST NAME: ALLEY

DATE/TIME OF TEST: 11/07/2019 12:36:34

Diagnostics Check		OK	12:28
Air Blank		0.000	12:29
Control Test		0.082	12:29
Air Blank		0.000	12:30
Subject Sample	# 1	0.197	12:30
Breath Volume		2.109	12:30
Air Blank		0.000	12:31
Air Blank		0.000	12:33
Subject Sample	# 2	0.207	12:34
Breath Volume		1.875	12:34
Air Blank		0.000	12:35
Control Test		0.080	12:35
Air Blank		0.000	12:36
Diagnostics Check		OK	12:36

Data Download Date/Time: 12/03/2019 16:40

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEORGE,SHANNON,N
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: LEZAMA,J
ARRESTING OFFICER AGENCY: FHP TROOP F
LAST AGENCY INSPECTION: 11/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: BAZYDLO

DATE/TIME OF TEST: 11/09/2019 03:01:18

Diagnostics Check		OK	02:54
Air Blank		0.000	02:55
Control Test		0.082	02:55
Air Blank		0.000	02:56
Subject Sample	# 1	0.168	02:56
Breath Volume		2.933	02:56
Air Blank		0.000	02:57
Air Blank		0.000	02:59
Subject Sample	# 2	0.179	02:59
Breath Volume		2.253	02:59
Air Blank		0.000	03:00
Control Test		0.081	03:00
Air Blank		0.000	03:01
Diagnostics Check		OK	03:01

Data Download Date/Time: 12/03/2019 16:40

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRUBIN,L,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: MAHOLTZ,S
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 11/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 15:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:32

SUBJECT'S LAST NAME: LEBER

DATE/TIME OF TEST: 11/10/2019 17:03:42

Diagnostics Check		OK	16:55
Air Blank		0.000	16:56
Control Test		0.082	16:56
Air Blank		0.000	16:56
Subject Sample	# 1	0.181	16:58
Breath Volume		2.121	16:58
Air Blank		0.000	16:59
Air Blank		0.000	17:00
Subject Sample	# 2	0.177	17:01
Breath Volume		1.382	17:01
Air Blank		0.000	17:02
Control Test		0.079	17:03
Air Blank		0.000	17:03
Diagnostics Check		OK	17:03

Data Download Date/Time: 12/03/2019 16:41

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACUNTO,ANGELINA,M
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: BIRTOLO,J
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 11/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:36

SUBJECT'S LAST NAME: KOHLHAGEN

DATE/TIME OF TEST: 11/11/2019 18:09:49

Diagnostics Check		OK	18:01
Air Blank		0.000	18:01
Control Test		0.082	18:02
Air Blank		0.000	18:02
Subject Sample	# 1	0.000	18:04
Breath Volume		2.343	18:04
Air Blank		0.000	18:05
Air Blank		0.000	18:07
Subject Sample	# 2	0.000	18:08
Breath Volume		1.242	18:08
Air Blank		0.000	18:08
Control Test		0.080	18:09
Air Blank		0.000	18:09
Diagnostics Check		OK	18:09

Data Download Date/Time: 12/03/2019 16:41

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,L
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: KENNEDY,DANIEL
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 11/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: PADRON

DATE/TIME OF TEST: 11/11/2019 23:03:43

Diagnostics Check		OK	22:57
Air Blank		0.000	22:57
Control Test		0.081	22:57
Air Blank		0.000	22:58
Subject Sample	# 1	0.244	22:59
Breath Volume		1.714	22:59
Air Blank		0.000	22:59
Air Blank		0.000	23:01
Subject Sample	# 2	0.240	23:02
Breath Volume		1.464	23:02
Air Blank		0.000	23:02
Control Test		0.080	23:03
Air Blank		0.000	23:03
Diagnostics Check		OK	23:03

Data Download Date/Time: 12/03/2019 16:41

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIERMAN,BRENT,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: HENDERSON,K
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 11/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: SWARTZ

DATE/TIME OF TEST: 11/12/2019 00:05:21

Diagnostics Check		OK	23:55
Air Blank		0.000	23:55
Control Test		0.081	23:55
Air Blank		0.000	23:56
Subject Sample	# 3	0.175	23:57
Breath Volume		1.773	23:57
Air Blank		0.000	23:58
Air Blank		0.000	23:59
Subject Sample	# 1	0.198	00:00
Breath Volume		2.785	00:00
Air Blank		0.000	00:01
Air Blank		0.000	00:02
Subject Sample	# 2	0.185	00:03
Breath Volume		1.460	00:03
Air Blank		0.000	00:04
Control Test		0.080	00:04
Air Blank		0.000	00:05
Diagnostics Check		OK	00:05

Data Download Date/Time: 12/03/2019 16:41

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRISCOLL,JAMES,P
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: DRISCOLL,JAMES
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 11/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: DAMASO

DATE/TIME OF TEST: 11/13/2019 03:25:05

Diagnostics Check		OK	03:21	
Air Blank		0.000	03:22	
Control Test		0.082	03:22	
Air Blank		0.000	03:22	
Subject Sample	# 1		03:23	Subject Test Refused
Breath Volume		0.000	03:23	
Air Blank		0.000	03:24	
Control Test		0.082	03:24	
Air Blank		0.000	03:24	
Diagnostics Check		OK	03:25	

Data Download Date/Time: 12/03/2019 16:41

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,L
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: KENNEDY,DANIEL
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 11/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:06

SUBJECT'S LAST NAME: MERRILL

DATE/TIME OF TEST: 11/15/2019 22:36:49

Diagnostics Check		OK	22:29
Air Blank		0.000	22:30
Control Test		0.081	22:30
Air Blank		0.000	22:30
Subject Sample	# 1	0.189	22:32
Breath Volume		1.968	22:32
Air Blank		0.000	22:32
Air Blank		0.000	22:34
Subject Sample	# 2	0.202	22:35
Breath Volume		2.882	22:35
Air Blank		0.000	22:35
Control Test		0.080	22:36
Air Blank		0.000	22:36
Diagnostics Check		OK	22:36

Data Download Date/Time: 12/03/2019 16:41

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKS,D,M
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: ZYNDA,T
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 11/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: SAINTIL

DATE/TIME OF TEST: 11/16/2019 04:56:52

Diagnostics Check		OK	04:45	
Air Blank		0.000	04:45	
Control Test		0.082	04:45	
Air Blank		0.000	04:46	
Subject Sample	# 1		04:49	No Sample Provided
Breath Volume		0.000	04:49	
Air Blank		0.000	04:49	
Air Blank		0.000	04:51	
Subject Sample	# 2	0.164	04:55	Volume Not Met
Breath Volume		0.082	04:55	
Air Blank		0.000	04:55	
Control Test		0.082	04:56	
Air Blank		0.000	04:56	
Diagnostics Check		OK	04:56	

Data Download Date/Time: 12/03/2019 16:41

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOEL,Y,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: MENDOZA,U
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 11/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 11/20/2019 00:13:18

Diagnostics Check		OK	23:58	
Air Blank		0.000	23:58	
Control Test		0.082	23:58	
Air Blank		0.000	23:59	
Subject Sample	# 1	0.150	00:00	
Breath Volume		1.132	00:00	
Air Blank		0.000	00:01	
Air Blank		0.000	00:03	
Subject Sample	# 2	0.123	00:06	Volume Not Met
Breath Volume		0.500	00:06	
Air Blank		0.000	00:07	
Air Blank		0.000	00:08	
Subject Sample	# 3		00:11	No Sample Provided
Breath Volume		0.000	00:11	
Air Blank		0.000	00:12	
Control Test		0.078	00:12	
Air Blank		0.000	00:13	
Diagnostics Check		OK	00:13	

Data Download Date/Time: 12/03/2019 16:41

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOEL,YVES,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: MENDOZA,U
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 11/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 11/20/2019 00:50:15

Diagnostics Check		OK	00:40	
Air Blank		0.000	00:40	
Control Test		0.081	00:40	
Air Blank		0.000	00:41	
Subject Sample	# 1	0.135	00:44	Volume Not Met
Breath Volume		0.308	00:44	
Air Blank		0.000	00:45	
Air Blank		0.000	00:46	
Subject Sample	# 2		00:48	Subject Test Refused
Breath Volume		0.160	00:48	
Air Blank		0.000	00:49	
Control Test		0.079	00:49	
Air Blank		0.000	00:50	
Diagnostics Check		OK	00:50	

Data Download Date/Time: 12/03/2019 16:41

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUGO,JONATHAN,B
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: REED,J
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 11/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: VASQUEZ VASQUEZ

DATE/TIME OF TEST: 11/20/2019 03:08:29

Diagnostics Check		OK	03:02
Air Blank		0.000	03:02
Control Test		0.080	03:03
Air Blank		0.000	03:03
Subject Sample	# 1	0.233	03:04
Breath Volume		1.488	03:04
Air Blank		0.000	03:04
Air Blank		0.000	03:06
Subject Sample	# 2	0.240	03:06
Breath Volume		1.476	03:06
Air Blank		0.000	03:07
Control Test		0.080	03:07
Air Blank		0.000	03:08
Diagnostics Check		OK	03:08

Data Download Date/Time: 12/03/2019 16:41

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,SAMANTHA,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: PEREZ,M
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 11/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: ERICKSON

DATE/TIME OF TEST: 11/23/2019 01:07:17

Diagnostics Check		OK	00:50	
Air Blank		0.000	00:51	
Control Test		0.082	00:51	
Air Blank		0.000	00:52	
Subject Sample	# 1	0.142	00:55	Slope Not Level
Breath Volume		1.167	00:55	
Air Blank		0.000	00:56	
Air Blank		0.000	00:57	
Subject Sample	# 2	0.164	01:01	
Breath Volume		1.585	01:01	
Air Blank		0.000	01:01	
Air Blank		0.000	01:03	
Subject Sample	# 3	0.162	01:05	
Breath Volume		1.777	01:05	
Air Blank		0.000	01:06	
Control Test		0.078	01:06	
Air Blank		0.000	01:07	
Diagnostics Check		OK	01:07	

Data Download Date/Time: 12/03/2019 16:41

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOPS,C,M
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: KAPPOCK,S
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 11/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: PETTO

DATE/TIME OF TEST: 11/23/2019 01:51:42

Diagnostics Check		OK	01:45
Air Blank		0.000	01:45
Control Test		0.080	01:45
Air Blank		0.000	01:46
Subject Sample	# 1	0.178	01:46
Breath Volume		1.785	01:46
Air Blank		0.000	01:47
Air Blank		0.000	01:49
Subject Sample	# 2	0.185	01:50
Breath Volume		1.359	01:50
Air Blank		0.000	01:50
Control Test		0.080	01:51
Air Blank		0.000	01:51
Diagnostics Check		OK	01:51

Data Download Date/Time: 12/03/2019 16:41

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NATSON,ANDRAE,M
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: STEVENSON,C
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 11/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: WHITLOCK

DATE/TIME OF TEST: 11/23/2019 03:36:54

Diagnostics Check		OK	03:30
Air Blank		0.000	03:30
Control Test		0.080	03:30
Air Blank		0.000	03:31
Subject Sample	# 1	0.194	03:31
Breath Volume		1.515	03:31
Air Blank		0.000	03:32
Air Blank		0.000	03:34
Subject Sample	# 2	0.192	03:35
Breath Volume		2.703	03:35
Air Blank		0.000	03:35
Control Test		0.081	03:36
Air Blank		0.000	03:36
Diagnostics Check		OK	03:36

Data Download Date/Time: 12/03/2019 16:41

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODI,DOMENIC,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: CORWIN,J
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 11/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: WHIDDON

DATE/TIME OF TEST: 11/23/2019 21:50:58

Diagnostics Check		OK	21:43
Air Blank		0.000	21:43
Control Test		0.081	21:43
Air Blank		0.000	21:44
Subject Sample	# 1	0.282	21:45
Breath Volume		1.550	21:45
Air Blank		0.000	21:45
Air Blank		0.000	21:47
Subject Sample	# 2	0.264	21:49
Breath Volume		1.796	21:49
Air Blank		0.000	21:49
Control Test		0.081	21:50
Air Blank		0.000	21:50
Diagnostics Check		OK	21:50

Data Download Date/Time: 12/03/2019 16:41

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,A,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: MANHOLTZ,S
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 11/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 15:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:06

SUBJECT'S LAST NAME: OMALLEY

DATE/TIME OF TEST: 11/24/2019 16:43:14

Diagnostics Check		OK	16:36
Air Blank		0.000	16:37
Control Test		0.082	16:37
Air Blank		0.000	16:38
Subject Sample	# 1	0.274	16:38
Breath Volume		1.402	16:38
Air Blank		0.000	16:39
Air Blank		0.000	16:41
Subject Sample	# 2	0.265	16:41
Breath Volume		1.777	16:41
Air Blank		0.000	16:42
Control Test		0.081	16:42
Air Blank		0.000	16:43
Diagnostics Check		OK	16:43

Data Download Date/Time: 12/03/2019 16:41

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TARR,L,J
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: LEZAMA,J
ARRESTING OFFICER AGENCY: FHP TROOP F
LAST AGENCY INSPECTION: 11/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: COMMERS

DATE/TIME OF TEST: 11/25/2019 01:06:51

Diagnostics Check		OK	01:00
Air Blank		0.000	01:00
Control Test		0.082	01:00
Air Blank		0.000	01:01
Subject Sample	# 1	0.000	01:02
Breath Volume		2.953	01:02
Air Blank		0.000	01:02
Air Blank		0.000	01:04
Subject Sample	# 2	0.000	01:05
Breath Volume		2.964	01:05
Air Blank		0.000	01:05
Control Test		0.080	01:06
Air Blank		0.000	01:06
Diagnostics Check		OK	01:06

Data Download Date/Time: 12/03/2019 16:41

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOPS,C,M
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: GROSSKLAS,D
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 11/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: LUCIO

DATE/TIME OF TEST: 11/25/2019 01:26:56

Diagnostics Check		OK	01:24	
Air Blank		0.000	01:24	
Control Test		0.082	01:24	
Air Blank		0.000	01:25	
Subject Sample	# 1		01:25	Subject Test Refused
Breath Volume		0.000	01:25	
Air Blank		0.000	01:26	
Control Test		0.081	01:26	
Air Blank		0.000	01:26	
Diagnostics Check		OK	01:26	

Data Download Date/Time: 12/03/2019 16:41

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRICE,M,D
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: GRANESE,J
ARRESTING OFFICER AGENCY: NAPLES PD
LAST AGENCY INSPECTION: 11/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: GULLETTE

DATE/TIME OF TEST: 11/26/2019 22:18:41

Diagnostics Check		OK	22:11
Air Blank		0.000	22:11
Control Test		0.082	22:11
Air Blank		0.000	22:12
Subject Sample	# 1	0.000	22:13
Breath Volume		3.179	22:13
Air Blank		0.000	22:14
Air Blank		0.000	22:16
Subject Sample	# 2	0.000	22:17
Breath Volume		3.289	22:17
Air Blank		0.000	22:17
Control Test		0.081	22:18
Air Blank		0.000	22:18
Diagnostics Check		OK	22:18

Data Download Date/Time: 12/03/2019 16:41

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAFFEE,C,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: DRUCKS,D
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 11/01/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: ISAACS

DATE/TIME OF TEST: 11/26/2019 22:42:19

Diagnostics Check		OK	22:35
Air Blank		0.000	22:36
Control Test		0.081	22:36
Air Blank		0.000	22:37
Subject Sample	# 1	0.000	22:37
Breath Volume		1.710	22:37
Air Blank		0.000	22:38
Air Blank		0.000	22:40
Subject Sample	# 2	0.000	22:40
Breath Volume		1.691	22:40
Air Blank		0.000	22:41
Control Test		0.080	22:41
Air Blank		0.000	22:42
Diagnostics Check		OK	22:42

Data Download Date/Time: 12/03/2019 16:41

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EMANS,ALEX,R
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: POUNDS,R
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 11/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: O'MEARA

DATE/TIME OF TEST: 11/26/2019 23:25:01

Diagnostics Check		OK	23:18
Air Blank		0.000	23:18
Control Test		0.081	23:19
Air Blank		0.000	23:19
Subject Sample	# 1	0.178	23:20
Breath Volume		1.578	23:20
Air Blank		0.000	23:20
Air Blank		0.000	23:22
Subject Sample	# 2	0.174	23:23
Breath Volume		1.445	23:23
Air Blank		0.000	23:24
Control Test		0.081	23:24
Air Blank		0.000	23:24
Diagnostics Check		OK	23:25

Data Download Date/Time: 12/03/2019 16:41

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVAS,W,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: LANTINI,N
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 11/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: GUZMANHERNANDEZ

DATE/TIME OF TEST: 11/28/2019 02:21:08

Diagnostics Check		OK	02:14
Air Blank		0.000	02:15
Control Test		0.082	02:15
Air Blank		0.000	02:16
Subject Sample	# 1	0.135	02:16
Breath Volume		1.953	02:16
Air Blank		0.000	02:17
Air Blank		0.000	02:19
Subject Sample	# 2	0.136	02:19
Breath Volume		3.707	02:19
Air Blank		0.000	02:20
Control Test		0.082	02:20
Air Blank		0.000	02:21
Diagnostics Check		OK	02:21

Data Download Date/Time: 12/03/2019 16:41

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,THOMAS,L
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: CORWIN,J
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 11/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: SIMS

DATE/TIME OF TEST: 11/29/2019 01:43:05

Diagnostics Check		OK	01:35
Air Blank		0.000	01:35
Control Test		0.082	01:36
Air Blank		0.000	01:36
Subject Sample	# 1	0.178	01:37
Breath Volume		1.167	01:37
Air Blank		0.000	01:37
Air Blank		0.000	01:39
Subject Sample	# 2	0.184	01:41
Breath Volume		1.347	01:41
Air Blank		0.000	01:42
Control Test		0.081	01:42
Air Blank		0.000	01:43
Diagnostics Check		OK	01:43

Data Download Date/Time: 12/03/2019 16:41

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORNELIUS,TIM,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: CORNELIUS,TIM
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 11/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:34

SUBJECT'S LAST NAME: SWADE

DATE/TIME OF TEST: 11/29/2019 19:01:45

Diagnostics Check		OK	18:55
Air Blank		0.000	18:55
Control Test		0.081	18:55
Air Blank		0.000	18:56
Subject Sample	# 1	0.258	18:56
Breath Volume		1.738	18:56
Air Blank		0.000	18:57
Air Blank		0.000	18:59
Subject Sample	# 2	0.260	19:00
Breath Volume		2.414	19:00
Air Blank		0.000	19:00
Control Test		0.080	19:01
Air Blank		0.000	19:01
Diagnostics Check		OK	19:01

Data Download Date/Time: 12/03/2019 16:41

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,MATTHEW,L
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: KORPOLINSKI,M
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 11/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: BEN ALEXANDER

DATE/TIME OF TEST: 11/29/2019 23:03:12

Diagnostics Check		OK	22:56
Air Blank		0.000	22:57
Control Test		0.082	22:57
Air Blank		0.000	22:58
Subject Sample	# 1	0.286	22:58
Breath Volume		2.246	22:58
Air Blank		0.000	22:59
Air Blank		0.000	23:00
Subject Sample	# 2	0.267	23:01
Breath Volume		3.503	23:01
Air Blank		0.000	23:02
Control Test		0.081	23:02
Air Blank		0.000	23:03
Diagnostics Check		OK	23:03

Data Download Date/Time: 12/03/2019 16:41

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANCILLA,RICHARD,P
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: LORENZO,ANNIEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 10/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: KEEGAN

DATE/TIME OF TEST: 11/01/2019 03:45:56

Diagnostics Check		OK	03:39
Air Blank		0.000	03:40
Control Test		0.079	03:40
Air Blank		0.000	03:40
Subject Sample	# 1	0.125	03:41
Breath Volume		3.515	03:41
Air Blank		0.000	03:42
Air Blank		0.000	03:43
Subject Sample	# 2	0.125	03:44
Breath Volume		3.636	03:44
Air Blank		0.000	03:44
Control Test		0.080	03:45
Air Blank		0.000	03:45
Diagnostics Check		OK	03:45

Data Download Date/Time: 11/03/2019 09:07

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNN,CDFC,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: ERICKSON,C
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 10/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: MILNE HESS

DATE/TIME OF TEST: 11/02/2019 00:10:40

Diagnostics Check		OK	00:03
Air Blank		0.000	00:03
Control Test		0.079	00:03
Air Blank		0.000	00:04
Subject Sample	# 1	0.244	00:05
Breath Volume		1.417	00:05
Air Blank		0.000	00:06
Air Blank		0.000	00:07
Subject Sample	# 2	0.243	00:08
Breath Volume		1.671	00:08
Air Blank		0.000	00:09
Control Test		0.082	00:10
Air Blank		0.000	00:10
Diagnostics Check		OK	00:10

Data Download Date/Time: 11/03/2019 09:07

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLATZ,KEVIN,L
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: LEUTHNER,NICK
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 10/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: MOODY

DATE/TIME OF TEST: 11/02/2019 05:06:42

Diagnostics Check		OK	04:59
Air Blank		0.000	05:00
Control Test		0.079	05:00
Air Blank		0.000	05:00
Subject Sample	# 1	0.151	05:01
Breath Volume		3.078	05:01
Air Blank		0.000	05:02
Air Blank		0.000	05:04
Subject Sample	# 2	0.137	05:05
Breath Volume		1.511	05:05
Air Blank		0.000	05:05
Control Test		0.079	05:06
Air Blank		0.000	05:06
Diagnostics Check		OK	05:06

Data Download Date/Time: 11/03/2019 09:07

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANCILLA,RICHARD,P
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: LORENZO,ANNIEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 10/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: WARD IV

DATE/TIME OF TEST: 11/06/2019 03:07:38

Diagnostics Check		OK	03:01
Air Blank		0.000	03:01
Control Test		0.080	03:01
Air Blank		0.000	03:02
Subject Sample	# 1	0.179	03:02
Breath Volume		3.839	03:02
Air Blank		0.000	03:03
Air Blank		0.000	03:05
Subject Sample	# 2	0.175	03:05
Breath Volume		4.218	03:05
Air Blank		0.000	03:06
Control Test		0.080	03:06
Air Blank		0.000	03:07
Diagnostics Check		OK	03:07

Data Download Date/Time: 12/03/2019 00:52

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORDSTROM,DEPUTY,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: COTE,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 10/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: ADMIN
ARREST TIME: 16:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:20

SUBJECT'S LAST NAME: KIRMO

DATE/TIME OF TEST: 11/07/2019 16:50:21

Diagnostics Check		OK	16:44
Air Blank		0.000	16:44
Control Test		0.079	16:44
Air Blank		0.000	16:45
Subject Sample	# 1	0.294	16:45
Breath Volume		2.542	16:45
Air Blank		0.000	16:46
Air Blank		0.000	16:48
Subject Sample	# 2	0.290	16:48
Breath Volume		1.839	16:48
Air Blank		0.000	16:49
Control Test		0.076	16:49
Air Blank		0.000	16:50
Diagnostics Check		OK	16:50

Data Download Date/Time: 12/03/2019 00:52

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JUDD,BOBBY,L
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: JUDD,BOBBY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 10/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: ADMIN
ARREST TIME: 21:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 11/07/2019 21:35:50

Diagnostics Check		OK	21:29
Air Blank		0.000	21:29
Control Test		0.079	21:29
Air Blank		0.000	21:30
Subject Sample	# 1	0.211	21:31
Breath Volume		2.945	21:31
Air Blank		0.000	21:31
Air Blank		0.000	21:33
Subject Sample	# 2	0.204	21:34
Breath Volume		2.785	21:34
Air Blank		0.000	21:34
Control Test		0.079	21:35
Air Blank		0.000	21:35
Diagnostics Check		OK	21:35

Data Download Date/Time: 12/03/2019 00:52

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANCILLA,RICHARD,P
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: LORENZO,ANNIEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 10/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: SCHOOLEY

DATE/TIME OF TEST: 11/10/2019 00:12:49

Diagnostics Check		OK	00:06
Air Blank		0.000	00:07
Control Test		0.079	00:07
Air Blank		0.000	00:08
Subject Sample	# 1	0.034	00:08
Breath Volume		2.964	00:08
Air Blank		0.000	00:09
Air Blank		0.000	00:10
Subject Sample	# 2	0.034	00:11
Breath Volume		3.574	00:11
Air Blank		0.000	00:11
Control Test		0.080	00:12
Air Blank		0.000	00:12
Diagnostics Check		OK	00:12

Data Download Date/Time: 12/03/2019 00:52

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZAMORA,DEP,D
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: COTE,DEP
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 10/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: SCHULTZ

DATE/TIME OF TEST: 11/10/2019 03:02:33

Diagnostics Check		OK	02:55
Air Blank		0.000	02:55
Control Test		0.080	02:56
Air Blank		0.000	02:56
Subject Sample	# 1	0.202	02:57
Breath Volume		2.414	02:57
Air Blank		0.000	02:58
Air Blank		0.000	03:00
Subject Sample	# 2	0.188	03:00
Breath Volume		1.437	03:00
Air Blank		0.000	03:01
Control Test		0.081	03:01
Air Blank		0.000	03:02
Diagnostics Check		OK	03:02

Data Download Date/Time: 12/03/2019 00:52

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZAMORA,DEP,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: REYES,DEP
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 10/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 11/11/2019 00:23:06

Diagnostics Check		OK	00:16
Air Blank		0.000	00:17
Control Test		0.079	00:17
Air Blank		0.000	00:18
Subject Sample	# 1	0.083	00:18
Breath Volume		1.953	00:18
Air Blank		0.000	00:19
Air Blank		0.000	00:21
Subject Sample	# 2	0.080	00:21
Breath Volume		1.898	00:21
Air Blank		0.000	00:22
Control Test		0.080	00:22
Air Blank		0.000	00:23
Diagnostics Check		OK	00:23

Data Download Date/Time: 12/03/2019 00:52

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEONARD,ALEYA,D
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: KNAPP,ERIC
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 10/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: GUNIA

DATE/TIME OF TEST: 11/11/2019 19:28:12

Diagnostics Check		OK	19:16	
Air Blank		0.000	19:16	
Control Test		0.079	19:17	
Air Blank		0.000	19:17	
Subject Sample	# 1	0.142	19:20	Volume Not Met
Breath Volume		0.343	19:20	
Air Blank		0.000	19:21	
Air Blank		0.000	19:23	
Subject Sample	# 2	0.137	19:26	Volume Not Met
Breath Volume		0.480	19:26	
Air Blank		0.000	19:27	
Control Test		0.080	19:27	
Air Blank		0.000	19:28	
Diagnostics Check		OK	19:28	

Data Download Date/Time: 12/03/2019 00:52

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JUDD,BOBBY,L
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: JUDD,BOBBY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 10/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: ADMIN
ARREST TIME: 20:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: GARON

DATE/TIME OF TEST: 11/12/2019 21:04:32

Diagnostics Check		OK	20:57
Air Blank		0.000	20:57
Control Test		0.078	20:58
Air Blank		0.000	20:58
Subject Sample	# 1	0.167	20:59
Breath Volume		2.964	20:59
Air Blank		0.000	21:00
Air Blank		0.000	21:01
Subject Sample	# 2	0.153	21:02
Breath Volume		2.191	21:02
Air Blank		0.000	21:03
Control Test		0.084	21:03
Air Blank		0.000	21:04
Diagnostics Check		OK	21:04

Data Download Date/Time: 12/03/2019 00:52

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JUDD,BOBBY,L
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: JUDD,BOBBY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 10/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: ADMIN
ARREST TIME: 00:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: GARON

DATE/TIME OF TEST: 11/13/2019 01:12:47

Diagnostics Check		OK	01:05
Air Blank		0.000	01:06
Control Test		0.080	01:06
Air Blank		0.000	01:07
Subject Sample	# 1	0.083	01:08
Breath Volume		2.957	01:08
Air Blank		0.000	01:08
Air Blank		0.000	01:10
Subject Sample	# 2	0.080	01:11
Breath Volume		3.550	01:11
Air Blank		0.000	01:11
Control Test		0.084	01:12
Air Blank		0.000	01:12
Diagnostics Check		OK	01:12

Data Download Date/Time: 12/03/2019 00:52

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JUDD,DEP,B
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: TAYLOR,R
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 10/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: DAVENPORT

DATE/TIME OF TEST: 11/13/2019 02:27:25

Diagnostics Check		OK	02:20
Air Blank		0.000	02:20
Control Test		0.079	02:20
Air Blank		0.000	02:21
Subject Sample	# 1	0.097	02:22
Breath Volume		1.519	02:22
Air Blank		0.000	02:23
Air Blank		0.000	02:25
Subject Sample	# 2	0.104	02:25
Breath Volume		1.503	02:25
Air Blank		0.000	02:26
Control Test		0.081	02:26
Air Blank		0.000	02:27
Diagnostics Check		OK	02:27

Data Download Date/Time: 12/03/2019 00:52

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JUDD,DEP,B
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: JUDD,BOBBY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 10/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: ADMIN
ARREST TIME: 04:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: GARON

DATE/TIME OF TEST: 11/13/2019 04:55:41

Diagnostics Check		OK	04:49
Air Blank		0.000	04:49
Control Test		0.079	04:50
Air Blank		0.000	04:50
Subject Sample	# 1	0.015	04:51
Breath Volume		2.808	04:51
Air Blank		0.000	04:51
Air Blank		0.000	04:53
Subject Sample	# 2	0.016	04:54
Breath Volume		2.929	04:54
Air Blank		0.000	04:54
Control Test		0.080	04:55
Air Blank		0.000	04:55
Diagnostics Check		OK	04:55

Data Download Date/Time: 12/03/2019 00:52

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEONARD,ALEYA,D
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: OWENS,NICHOLAS
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 10/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: SANDY

DATE/TIME OF TEST: 11/15/2019 22:32:38

Diagnostics Check		OK	22:25
Air Blank		0.000	22:26
Control Test		0.079	22:26
Air Blank		0.000	22:27
Subject Sample	# 1	0.223	22:27
Breath Volume		1.871	22:27
Air Blank		0.000	22:28
Air Blank		0.000	22:30
Subject Sample	# 2	0.214	22:30
Breath Volume		1.339	22:30
Air Blank		0.000	22:31
Control Test		0.080	22:31
Air Blank		0.000	22:32
Diagnostics Check		OK	22:32

Data Download Date/Time: 12/03/2019 00:52

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEONARD,ALEYA,D
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: DAVIDSON,MICHAEL
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 10/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: GIZZI

DATE/TIME OF TEST: 11/16/2019 02:17:47

Diagnostics Check		OK	02:11
Air Blank		0.000	02:11
Control Test		0.080	02:11
Air Blank		0.000	02:12
Subject Sample	# 1	0.221	02:13
Breath Volume		1.671	02:13
Air Blank		0.000	02:13
Air Blank		0.000	02:15
Subject Sample	# 2	0.239	02:16
Breath Volume		2.085	02:16
Air Blank		0.000	02:16
Control Test		0.079	02:17
Air Blank		0.000	02:17
Diagnostics Check		OK	02:17

Data Download Date/Time: 12/03/2019 00:52

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEONARD,ALEYA,D
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: LEONARD,ALEYA
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 10/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: ADMIN
ARREST TIME: 00:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: LEONARD

DATE/TIME OF TEST: 11/16/2019 02:31:21

Diagnostics Check	OK	02:29	
Air Blank	0.000	02:30	
Control Test	0.080	02:30	
Air Blank		02:30	RFI Detect
Air Blank	0.000	02:31	

Data Download Date/Time: 12/03/2019 00:52

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEUTHNER,N.,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: LEUTHNER,N.
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 10/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 11/16/2019 03:52:56

Diagnostics Check		OK	03:45
Air Blank		0.000	03:46
Control Test		0.080	03:46
Air Blank		0.000	03:46
Subject Sample	# 1	0.216	03:48
Breath Volume		1.816	03:48
Air Blank		0.000	03:48
Air Blank		0.000	03:50
Subject Sample	# 2	0.205	03:51
Breath Volume		2.093	03:51
Air Blank		0.000	03:51
Control Test		0.080	03:52
Air Blank		0.000	03:52
Diagnostics Check		OK	03:52

Data Download Date/Time: 12/03/2019 00:52

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JUDD,DEP,B
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: DAVIDSON,MIKE
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 10/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: RODRIGUEZ CHAVEZ

DATE/TIME OF TEST: 11/17/2019 04:56:54

Diagnostics Check		OK	04:49
Air Blank		0.000	04:49
Control Test		0.080	04:50
Air Blank		0.000	04:50
Subject Sample	# 1	0.152	04:51
Breath Volume		1.898	04:51
Air Blank		0.000	04:52
Air Blank		0.000	04:54
Subject Sample	# 2	0.162	04:55
Breath Volume		1.375	04:55
Air Blank		0.000	04:55
Control Test		0.081	04:56
Air Blank		0.000	04:56
Diagnostics Check		OK	04:56

Data Download Date/Time: 12/03/2019 00:52

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,CDFC,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: PAPPAS,A
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 10/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: CANALES

DATE/TIME OF TEST: 11/17/2019 22:25:14

Diagnostics Check		OK	22:17
Air Blank		0.000	22:17
Control Test		0.081	22:18
Air Blank		0.000	22:18
Subject Sample	# 1	0.000	22:19
Breath Volume		2.007	22:19
Air Blank		0.000	22:20
Air Blank		0.000	22:22
Subject Sample	# 2	0.000	22:23
Breath Volume		2.054	22:23
Air Blank		0.000	22:24
Control Test		0.081	22:24
Air Blank		0.000	22:25
Diagnostics Check		OK	22:25

Data Download Date/Time: 12/03/2019 00:52

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,CDFC,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: TAYLOR,R
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 10/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: SWIKA POST

DATE/TIME OF TEST: 11/18/2019 03:04:30

Diagnostics Check		OK	02:53	
Air Blank		0.000	02:54	
Control Test		0.080	02:54	
Air Blank		0.000	02:54	
Subject Sample	# 1	0.217	02:57	
Breath Volume		1.488	02:57	
Air Blank		0.000	02:58	
Air Blank		0.000	02:59	
Subject Sample	# 2	0.184	03:02	Volume Not Met
Breath Volume		0.027	03:02	
Air Blank		0.000	03:03	
Air Blank		0.000	03:04	

Data Download Date/Time: 12/03/2019 00:52

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RONGA,DEPUTY,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: HENSHAW,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 10/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 12:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:16

SUBJECT'S LAST NAME: BADALAMENTI

DATE/TIME OF TEST: 11/20/2019 18:02:13

Diagnostics Check		OK	17:50	
Air Blank		0.000	17:50	
Control Test		0.079	17:51	
Air Blank		0.000	17:51	
Subject Sample	# 1	0.122	17:54	Volume Not Met
Breath Volume		1.093	17:54	
Air Blank		0.000	17:55	
Air Blank		0.000	17:57	
Subject Sample	# 2	0.125	18:00	Volume Not Met
Breath Volume		0.128	18:00	
Air Blank		0.000	18:01	
Control Test		0.080	18:01	
Air Blank		0.000	18:02	
Diagnostics Check		OK	18:02	

Data Download Date/Time: 12/03/2019 00:52

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JUDD,BOBBY,L
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: PAPPAS,A
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 10/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: CUTRELL

DATE/TIME OF TEST: 11/21/2019 03:53:23

Diagnostics Check		OK	03:46
Air Blank		0.000	03:46
Control Test		0.080	03:47
Air Blank		0.000	03:47
Subject Sample	# 1	0.205	03:48
Breath Volume		4.539	03:48
Air Blank		0.000	03:49
Air Blank		0.000	03:50
Subject Sample	# 2	0.198	03:51
Breath Volume		4.070	03:51
Air Blank		0.000	03:52
Control Test		0.082	03:52
Air Blank		0.000	03:53
Diagnostics Check		OK	03:53

Data Download Date/Time: 12/03/2019 00:52

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JUDD,DEP,B
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: DAVIDSON,MIKE
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 10/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: BLOOM

DATE/TIME OF TEST: 11/22/2019 00:03:58

Diagnostics Check		OK	23:56
Air Blank		0.000	23:57
Control Test		0.079	23:57
Air Blank		0.000	23:58
Subject Sample	# 2	0.141	23:58
Breath Volume		3.515	23:58
Air Blank		0.000	23:59
Air Blank		0.000	00:01
Subject Sample	# 1	0.141	00:02
Breath Volume		1.773	00:02
Air Blank		0.000	00:03
Control Test		0.082	00:03
Air Blank		0.000	00:03
Diagnostics Check		OK	00:03

Data Download Date/Time: 12/03/2019 00:52

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANCILLA,RICHARD,P
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: ROBLES,NICHOLAS
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 10/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: SCHWEDE

DATE/TIME OF TEST: 11/22/2019 19:56:27

Diagnostics Check		OK	19:50
Air Blank		0.000	19:50
Control Test		0.079	19:51
Air Blank		0.000	19:51
Subject Sample	# 1	0.305	19:51
Breath Volume		3.847	19:51
Air Blank		0.000	19:52
Air Blank		0.000	19:54
Subject Sample	# 2	0.307	19:54
Breath Volume		3.863	19:54
Air Blank		0.000	19:55
Control Test		0.080	19:55
Air Blank		0.000	19:56
Diagnostics Check		OK	19:56

Data Download Date/Time: 12/03/2019 00:52

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZAMORA,DEP,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: ANDERSON,DEP
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 10/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: MINGS

DATE/TIME OF TEST: 11/23/2019 02:57:43

Diagnostics Check		OK	02:46	
Air Blank		0.000	02:46	
Control Test		0.079	02:47	
Air Blank		0.000	02:47	
Subject Sample	# 1	0.324	02:50	Volume Not Met
Breath Volume		0.101	02:50	
Air Blank		0.000	02:51	
Air Blank		0.000	02:53	
Subject Sample	# 2	0.264	02:56	Volume Not Met
Breath Volume		0.496	02:56	
Air Blank			02:57	Purge Fail
Air Blank		0.000	02:57	

Data Download Date/Time: 12/03/2019 00:52

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZAMORA,DEP,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: IERFINO,DEP
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 10/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: GRANT

DATE/TIME OF TEST: 11/23/2019 23:40:55

Diagnostics Check		OK	23:37	
Air Blank		0.000	23:37	
Control Test		0.077	23:38	
Air Blank		0.000	23:38	
Subject Sample	# 1		23:40	Sequence Aborted
Breath Volume		1.078	23:40	
Air Blank		0.000	23:40	

Data Download Date/Time: 12/03/2019 00:52

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNN,CDFC,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: DUNN,CDFC
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 10/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: ADMIN
ARREST TIME: 14:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:40

SUBJECT'S LAST NAME: DOE

DATE/TIME OF TEST: 11/25/2019 15:19:37

Diagnostics Check		OK	15:13
Air Blank		0.000	15:13
Control Test		0.080	15:13
Air Blank		0.000	15:14
Subject Sample	# 1	0.000	15:15
Breath Volume		3.078	15:15
Air Blank		0.000	15:15
Air Blank		0.000	15:17
Subject Sample	# 2	0.000	15:18
Breath Volume		2.519	15:18
Air Blank		0.000	15:18
Control Test		0.082	15:18
Air Blank		0.000	15:19
Diagnostics Check		OK	15:19

Data Download Date/Time: 12/03/2019 00:52

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANCILLA,RICHARD,P
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: HARTMANSTORFER,CHRISTOPHER
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 10/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: LYNCH JR

DATE/TIME OF TEST: 11/27/2019 23:55:05

Diagnostics Check		OK	23:48
Air Blank		0.000	23:49
Control Test		0.080	23:49
Air Blank		0.000	23:49
Subject Sample	# 1	0.128	23:50
Breath Volume		2.609	23:50
Air Blank		0.000	23:51
Air Blank		0.000	23:52
Subject Sample	# 2	0.121	23:53
Breath Volume		3.000	23:53
Air Blank		0.000	23:54
Control Test		0.081	23:54
Air Blank		0.000	23:55
Diagnostics Check		OK	23:55

Data Download Date/Time: 12/03/2019 00:52

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEONARD,ALEYA,D
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: DAVIDSON,MICHAEL
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 10/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: CORBIN

DATE/TIME OF TEST: 11/29/2019 23:47:18

Diagnostics Check		OK	23:39
Air Blank		0.000	23:39
Control Test		0.079	23:39
Air Blank		0.000	23:40
Subject Sample	# 1	0.100	23:41
Breath Volume		1.390	23:41
Air Blank		0.000	23:42
Air Blank		0.000	23:44
Subject Sample	# 2	0.097	23:45
Breath Volume		1.210	23:45
Air Blank		0.000	23:46
Control Test		0.081	23:46
Air Blank		0.000	23:47
Diagnostics Check		OK	23:47

Data Download Date/Time: 12/03/2019 00:53

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEONARD,ALEYA,D
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: LEUTHNER,NICHOLAS
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 10/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:36

SUBJECT'S LAST NAME: OSBORNE

DATE/TIME OF TEST: 11/30/2019 05:04:31

Diagnostics Check		OK	04:57
Air Blank		0.000	04:57
Control Test		0.079	04:57
Air Blank		0.000	04:58
Subject Sample	# 1	0.220	04:59
Breath Volume		2.140	04:59
Air Blank		0.000	04:59
Air Blank		0.000	05:01
Subject Sample	# 2	0.201	05:02
Breath Volume		1.109	05:02
Air Blank		0.000	05:03
Control Test		0.080	05:03
Air Blank		0.000	05:04
Diagnostics Check		OK	05:04

Data Download Date/Time: 12/03/2019 00:53

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEACH,ZACHARY,M
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: TOMASEGOVICH,KODY
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: DEAN

DATE/TIME OF TEST: 11/01/2019 02:53:34

Diagnostics Check		OK	02:47	
Air Blank		0.000	02:48	
Control Test		0.081	02:48	
Air Blank		0.000	02:49	
Subject Sample	# 1	0.132	02:49	
Breath Volume		4.648	02:49	
Air Blank		0.000	02:50	
Air Blank		0.000	02:52	
Subject Sample	# 2	0.129	02:52	
Breath Volume		4.523	02:52	
Air Blank			02:53	RFI Detect
Air Blank		0.000	02:53	

Data Download Date/Time: 11/08/2019 19:49

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINDBLOM,ROBERT,A
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: COLLINS,CHASE
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 12:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:34

SUBJECT'S LAST NAME: WINERICH

DATE/TIME OF TEST: 11/01/2019 13:02:53

Diagnostics Check		OK	12:56	
Air Blank		0.000	12:57	
Control Test		0.079	12:57	
Air Blank		0.000	12:58	
Subject Sample	# 1	0.000	12:58	
Breath Volume		3.000	12:58	
Air Blank		0.000	12:58	
Air Blank		0.000	13:00	
Subject Sample	# 2	0.000	13:01	
Breath Volume		3.089	13:01	
Air Blank		0.000	13:01	
Control Test		0.079	13:02	
Air Blank			13:02	RFI Detect
Air Blank		0.000	13:02	

Data Download Date/Time: 11/08/2019 19:49

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALEY,TYLER,J
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: CALEY,TYLER
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: NOLAN

DATE/TIME OF TEST: 11/01/2019 20:21:53

Diagnostics Check		OK	20:11
Air Blank		0.000	20:12
Control Test		0.081	20:12
Air Blank		0.000	20:12
Subject Sample	# 1	0.094	20:15
Breath Volume		1.562	20:15
Air Blank		0.000	20:15
Air Blank		0.000	20:17
Subject Sample	# 2	0.092	20:20
Breath Volume		1.484	20:20
Air Blank		0.000	20:20
Control Test		0.078	20:21
Air Blank		0.000	20:21
Diagnostics Check		OK	20:21

Data Download Date/Time: 11/08/2019 19:49

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUZBEE,HEATHER,N
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: WILLIAMS,BRANDON
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:51

SUBJECT'S LAST NAME: HUDGINS

DATE/TIME OF TEST: 11/01/2019 22:25:06

Diagnostics Check		OK	22:18
Air Blank		0.000	22:19
Control Test		0.079	22:19
Air Blank		0.000	22:19
Subject Sample	# 1	0.152	22:20
Breath Volume		2.824	22:20
Air Blank		0.000	22:21
Air Blank		0.000	22:23
Subject Sample	# 2	0.152	22:23
Breath Volume		2.921	22:23
Air Blank		0.000	22:24
Control Test		0.079	22:24
Air Blank		0.000	22:25
Diagnostics Check		OK	22:25

Data Download Date/Time: 11/08/2019 19:49

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GENTRY,BENJAMIN,W
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: GENTRY,BENJAMIN
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:50

SUBJECT'S LAST NAME: LETT

DATE/TIME OF TEST: 11/03/2019 06:18:14

Diagnostics Check		OK	06:10
Air Blank		0.000	06:10
Control Test		0.081	06:11
Air Blank		0.000	06:11
Subject Sample	# 1	0.000	06:13
Breath Volume		1.757	06:13
Air Blank		0.000	06:13
Air Blank		0.000	06:15
Subject Sample	# 2	0.000	06:16
Breath Volume		2.105	06:16
Air Blank		0.000	06:17
Control Test		0.078	06:17
Air Blank		0.000	06:18
Diagnostics Check		OK	06:18

Data Download Date/Time: 11/08/2019 19:49

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,AMANDA,L
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: DAVIS,AMANDA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: BRADLEY

DATE/TIME OF TEST: 11/07/2019 23:57:11

Diagnostics Check		OK	23:50
Air Blank		0.000	23:50
Control Test		0.079	23:51
Air Blank		0.000	23:51
Subject Sample	# 1	0.178	23:52
Breath Volume		3.394	23:52
Air Blank		0.000	23:53
Air Blank		0.000	23:55
Subject Sample	# 2	0.183	23:55
Breath Volume		4.363	23:55
Air Blank		0.000	23:56
Control Test		0.078	23:56
Air Blank		0.000	23:57
Diagnostics Check		OK	23:57

Data Download Date/Time: 11/08/2019 19:49

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,JAMIE,L
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: NELSON,JAMIE
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 11/08/2019 01:49:06

Diagnostics Check		OK	01:42
Air Blank		0.000	01:43
Control Test		0.078	01:43
Air Blank		0.000	01:44
Subject Sample	# 1	0.215	01:44
Breath Volume		3.558	01:44
Air Blank		0.000	01:45
Air Blank		0.000	01:47
Subject Sample	# 2	0.199	01:47
Breath Volume		2.679	01:47
Air Blank		0.000	01:48
Control Test		0.078	01:48
Air Blank		0.000	01:49
Diagnostics Check		OK	01:49

Data Download Date/Time: 11/08/2019 19:49

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEARN,SHELDON,J
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: HEARN,SHELDON
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 11/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: KUSWANDY

DATE/TIME OF TEST: 11/10/2019 22:42:59

Diagnostics Check		OK	22:31	
Air Blank		0.000	22:31	
Control Test		0.080	22:32	
Air Blank		0.000	22:32	
Subject Sample	# 1	0.127	22:35	Volume Not Met
Breath Volume		0.046	22:35	
Air Blank		0.000	22:36	
Air Blank		0.000	22:38	
Subject Sample	# 2	0.179	22:41	Volume Not Met
Breath Volume		0.113	22:41	
Air Blank		0.000	22:42	
Control Test		0.078	22:42	
Air Blank		0.000	22:42	
Diagnostics Check		OK	22:42	

Data Download Date/Time: 12/12/2019 11:24

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,AMANDA,L
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: DAVIS,AMANDA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 11/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:23

SUBJECT'S LAST NAME: MCCRAY

DATE/TIME OF TEST: 11/12/2019 22:50:29

Diagnostics Check		OK	22:44
Air Blank		0.000	22:44
Control Test		0.081	22:45
Air Blank		0.000	22:45
Subject Sample	# 1	0.074	22:46
Breath Volume		1.476	22:46
Air Blank		0.000	22:46
Air Blank		0.000	22:48
Subject Sample	# 2	0.075	22:48
Breath Volume		1.425	22:48
Air Blank		0.000	22:49
Control Test		0.081	22:49
Air Blank		0.000	22:50
Diagnostics Check		OK	22:50

Data Download Date/Time: 12/12/2019 11:24

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUDSON,JOSHUA,K
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: HUDSON,JOSHUA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 11/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 11/15/2019 23:45:51

Diagnostics Check		OK	23:42	
Air Blank		0.000	23:42	
Control Test		0.079	23:42	
Air Blank		0.000	23:43	
Subject Sample	# 1		23:44	Subject Test Refused
Breath Volume		0.000	23:44	
Air Blank		0.000	23:44	
Control Test		0.078	23:45	
Air Blank		0.000	23:45	
Diagnostics Check		OK	23:45	

Data Download Date/Time: 12/12/2019 11:24

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKIPPER,KYLE,A
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: SKIPPER,KYLE
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 11/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: RIMMER

DATE/TIME OF TEST: 11/16/2019 00:41:39

Diagnostics Check		OK	00:35
Air Blank		0.000	00:35
Control Test		0.080	00:36
Air Blank		0.000	00:36
Subject Sample	# 1	0.165	00:37
Breath Volume		2.113	00:37
Air Blank		0.000	00:37
Air Blank		0.000	00:39
Subject Sample	# 2	0.163	00:40
Breath Volume		2.382	00:40
Air Blank		0.000	00:40
Control Test		0.079	00:41
Air Blank		0.000	00:41
Diagnostics Check		OK	00:41

Data Download Date/Time: 12/12/2019 11:24

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,JAMIE,L
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: WILLIAMS,BRANDON
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 11/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: HOGUE

DATE/TIME OF TEST: 11/16/2019 01:02:18

Diagnostics Check		OK	00:56
Air Blank		0.000	00:56
Control Test		0.079	00:57
Air Blank		0.000	00:57
Subject Sample	# 1	0.000	00:58
Breath Volume		3.777	00:58
Air Blank		0.000	00:58
Air Blank		0.000	01:00
Subject Sample	# 2	0.000	01:00
Breath Volume		2.082	01:00
Air Blank		0.000	01:01
Control Test		0.079	01:01
Air Blank		0.000	01:02
Diagnostics Check		OK	01:02

Data Download Date/Time: 12/12/2019 11:24

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,JAMIE,L
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: BROWN,BRENDAN
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 11/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 12:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: FAYARD

DATE/TIME OF TEST: 11/16/2019 01:28:46

Diagnostics Check		OK	01:22
Air Blank		0.000	01:23
Control Test		0.080	01:23
Air Blank		0.000	01:23
Subject Sample	# 1	0.026	01:24
Breath Volume		4.820	01:24
Air Blank		0.000	01:24
Air Blank		0.000	01:26
Subject Sample	# 2	0.026	01:27
Breath Volume		4.515	01:27
Air Blank		0.000	01:27
Control Test		0.080	01:28
Air Blank		0.000	01:28
Diagnostics Check		OK	01:28

Data Download Date/Time: 12/12/2019 11:24

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOURNEY,KEITH,M
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: TOURNEY,KEITH
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 11/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: SIMS

DATE/TIME OF TEST: 11/16/2019 02:22:28

Diagnostics Check		OK	02:16
Air Blank		0.000	02:16
Control Test		0.081	02:16
Air Blank		0.000	02:17
Subject Sample	# 1	0.146	02:17
Breath Volume		1.558	02:17
Air Blank		0.000	02:18
Air Blank		0.000	02:20
Subject Sample	# 2	0.151	02:20
Breath Volume		2.316	02:20
Air Blank		0.000	02:21
Control Test		0.079	02:21
Air Blank		0.000	02:22
Diagnostics Check		OK	02:22

Data Download Date/Time: 12/12/2019 11:24

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANDRY,KATHERINE,E
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: LANDRY,KATHERINE
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 11/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A1
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: COLEMAN

DATE/TIME OF TEST: 11/16/2019 23:27:31

Diagnostics Check		OK	23:20
Air Blank		0.000	23:21
Control Test		0.081	23:21
Air Blank		0.000	23:22
Subject Sample	# 1	0.056	23:22
Breath Volume		1.769	23:22
Air Blank		0.000	23:23
Air Blank		0.000	23:25
Subject Sample	# 2	0.052	23:26
Breath Volume		1.929	23:26
Air Blank		0.000	23:26
Control Test		0.080	23:26
Air Blank		0.000	23:27
Diagnostics Check		OK	23:27

Data Download Date/Time: 12/12/2019 11:24

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TASCHETTA,JOSEPH,S
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: TASCHETTA,JOSEPH
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 11/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A1
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: RENEAU

DATE/TIME OF TEST: 11/18/2019 03:47:47

Diagnostics Check		OK	03:32	
Air Blank		0.000	03:32	
Control Test		0.080	03:33	
Air Blank		0.000	03:33	
Subject Sample	# 1	0.169	03:35	
Breath Volume		1.304	03:35	
Air Blank		0.000	03:36	
Air Blank		0.000	03:38	
Subject Sample	# 2	0.167	03:41	Volume Not Met
Breath Volume		0.015	03:41	
Air Blank		0.000	03:42	
Air Blank		0.000	03:43	
Subject Sample	# 3	0.158	03:46	
Breath Volume		3.628	03:46	
Air Blank		0.000	03:46	
Control Test		0.077	03:47	
Air Blank		0.000	03:47	
Diagnostics Check		OK	03:47	

Data Download Date/Time: 12/12/2019 11:24

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,AMANDA,L
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: DAVIS,AMANDA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 11/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A1
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 11/20/2019 22:08:40

Diagnostics Check		OK	22:01
Air Blank		0.000	22:01
Control Test		0.080	22:01
Air Blank		0.000	22:02
Subject Sample	# 1	0.080	22:03
Breath Volume		1.875	22:03
Air Blank		0.000	22:04
Air Blank		0.000	22:06
Subject Sample	# 2	0.071	22:07
Breath Volume		1.871	22:07
Air Blank		0.000	22:07
Control Test		0.080	22:08
Air Blank		0.000	22:08
Diagnostics Check		OK	22:08

Data Download Date/Time: 12/12/2019 11:24

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALEY,TYLER,J
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: WILLIAMS,BRANDON
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 11/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A1
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: LOY

DATE/TIME OF TEST: 11/22/2019 01:43:31

Diagnostics Check		OK	01:36
Air Blank		0.000	01:36
Control Test		0.080	01:37
Air Blank		0.000	01:37
Subject Sample	# 1	0.114	01:38
Breath Volume		1.574	01:38
Air Blank		0.000	01:39
Air Blank		0.000	01:41
Subject Sample	# 2	0.113	01:41
Breath Volume		1.808	01:41
Air Blank		0.000	01:42
Control Test		0.078	01:42
Air Blank		0.000	01:43
Diagnostics Check		OK	01:43

Data Download Date/Time: 12/12/2019 11:24

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BYRD,BRANDON,C
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: BYRD,BRANDON
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 11/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A1
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: GONZALEZ-GARCIA

DATE/TIME OF TEST: 11/23/2019 00:02:36

Diagnostics Check		OK	23:50	
Air Blank		0.000	23:51	
Control Test		0.079	23:51	
Air Blank		0.000	23:52	
Subject Sample	# 2	0.242	23:55	Volume Not Met
Breath Volume		0.031	23:55	
Air Blank		0.000	23:56	
Air Blank		0.000	23:57	
Subject Sample	# 1	0.205	00:00	Volume Not Met
Breath Volume		0.390	00:00	
Air Blank		0.000	00:01	
Control Test		0.077	00:01	
Air Blank		0.000	00:02	
Diagnostics Check		OK	00:02	

Data Download Date/Time: 12/12/2019 11:24

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BYRD,BRANDON,C
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: BYRD,BRANDON
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 11/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A1
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: GONZALEZ-GARCIA

DATE/TIME OF TEST: 11/23/2019 00:16:34

Diagnostics Check		OK	00:04	
Air Blank		0.000	00:05	
Control Test		0.076	00:05	
Air Blank		0.000	00:06	
Subject Sample	# 1	0.156	00:09	Volume Not Met
Breath Volume		0.054	00:09	
Air Blank		0.000	00:10	
Air Blank		0.000	00:11	
Subject Sample	# 2	0.170	00:14	Volume Not Met
Breath Volume		0.039	00:14	
Air Blank		0.000	00:15	
Control Test		0.076	00:15	
Air Blank		0.000	00:16	
Diagnostics Check		OK	00:16	

Data Download Date/Time: 12/12/2019 11:24

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANDRY,KATHERINE,E
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: LANDRY,KATHERINE
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 11/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A1
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: JAMES

DATE/TIME OF TEST: 11/30/2019 03:59:25

Diagnostics Check		OK	03:54	
Air Blank		0.000	03:55	
Control Test		0.080	03:55	
Air Blank		0.000	03:56	
Subject Sample	# 1		03:57	RFI Detect
Breath Volume		1.332	03:57	
Air Blank		0.000	03:58	
Control Test		0.079	03:58	
Air Blank		0.000	03:59	
Diagnostics Check		OK	03:59	

Data Download Date/Time: 12/12/2019 11:24

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANDRY,KATHERINE,E
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: LANDRY,KATHERINE
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 11/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A1
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: JAMES

DATE/TIME OF TEST: 11/30/2019 04:09:04

Diagnostics Check		OK	04:02
Air Blank		0.000	04:02
Control Test		0.078	04:02
Air Blank		0.000	04:03
Subject Sample	# 1	0.188	04:03
Breath Volume		1.476	04:03
Air Blank		0.000	04:04
Air Blank		0.000	04:06
Subject Sample	# 2	0.182	04:07
Breath Volume		1.375	04:07
Air Blank		0.000	04:08
Control Test		0.079	04:08
Air Blank		0.000	04:08
Diagnostics Check		OK	04:09

Data Download Date/Time: 12/12/2019 11:24

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RESTIFO,CHARLES,T
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: RESTIFO,CHARLES
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 11/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A1
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:36

SUBJECT'S LAST NAME: EXANTUS

DATE/TIME OF TEST: 11/30/2019 19:08:12

Diagnostics Check		OK	18:59
Air Blank		0.000	19:00
Control Test		0.079	19:00
Air Blank		0.000	19:01
Subject Sample	# 1	0.211	19:03
Breath Volume		2.121	19:03
Air Blank		0.000	19:04
Air Blank		0.000	19:06
Subject Sample	# 2	0.212	19:06
Breath Volume		1.261	19:06
Air Blank		0.000	19:07
Control Test		0.077	19:07
Air Blank		0.000	19:08
Diagnostics Check		OK	19:08

Data Download Date/Time: 12/12/2019 11:24

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUDSON,ANDREW,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: HUDSON,ANDREW
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 04:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: AROCHO

DATE/TIME OF TEST: 11/01/2019 05:06:36

Diagnostics Check		OK	05:00
Air Blank		0.000	05:00
Control Test		0.080	05:00
Air Blank		0.000	05:01
Subject Sample	# 1	0.046	05:02
Breath Volume		3.925	05:02
Air Blank		0.000	05:02
Air Blank		0.000	05:04
Subject Sample	# 2	0.044	05:05
Breath Volume		2.949	05:05
Air Blank		0.000	05:05
Control Test		0.079	05:06
Air Blank		0.000	05:06
Diagnostics Check		OK	05:06

Data Download Date/Time: 11/21/2019 12:26

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: CHLEBANOWSKI,RYAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: HUTCHINSON

DATE/TIME OF TEST: 11/02/2019 04:23:16

Diagnostics Check		OK	04:20	
Air Blank		0.000	04:20	
Control Test		0.080	04:21	
Air Blank		0.000	04:21	
Subject Sample	# 1		04:21	Subject Test Refused
Breath Volume		0.000	04:21	
Air Blank		0.000	04:22	
Control Test		0.079	04:22	
Air Blank		0.000	04:23	
Diagnostics Check		OK	04:23	

Data Download Date/Time: 11/21/2019 12:26

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOGERELEY,NEIL,F
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: MOGERLEY,NEIL
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 18:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 11/04/2019 18:50:26

Diagnostics Check		OK	18:47	
Air Blank		0.000	18:47	
Control Test		0.080	18:48	
Air Blank		0.000	18:48	
Subject Sample	# 1		18:49	Subject Test Refused
Breath Volume		0.000	18:49	
Air Blank		0.000	18:49	
Control Test		0.080	18:49	
Air Blank		0.000	18:50	
Diagnostics Check		OK	18:50	

Data Download Date/Time: 11/21/2019 12:26

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOGERLEY,NEIL,F
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: MOGERLEY,NEIL
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 21:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 11/05/2019 22:21:48

Diagnostics Check		OK	22:18	
Air Blank		0.000	22:18	
Control Test		0.080	22:18	
Air Blank		0.000	22:19	
Subject Sample	# 1		22:20	Subject Test Refused
Breath Volume		0.000	22:20	
Air Blank		0.000	22:20	
Control Test		0.080	22:21	
Air Blank		0.000	22:21	
Diagnostics Check		OK	22:21	

Data Download Date/Time: 11/21/2019 12:26

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TYLER,BRYAN,D
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: OGGERI,TIM
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: MCALISTER

DATE/TIME OF TEST: 11/07/2019 01:59:28

Diagnostics Check		OK	01:55	
Air Blank		0.000	01:55	
Control Test		0.079	01:55	
Air Blank		0.000	01:56	
Subject Sample	# 1		01:57	RFI Detect
Breath Volume		0.933	01:57	
Air Blank		0.000	01:58	
Control Test		0.079	01:58	
Air Blank		0.000	01:59	
Diagnostics Check		OK	01:59	

Data Download Date/Time: 11/21/2019 12:26

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TYLER,BRYAN,D
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: OGGERI,TIM
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: MCALISTER

DATE/TIME OF TEST: 11/07/2019 02:08:09

Diagnostics Check		OK	02:01
Air Blank		0.000	02:01
Control Test		0.078	02:02
Air Blank		0.000	02:02
Subject Sample	# 1	0.272	02:03
Breath Volume		1.800	02:03
Air Blank		0.000	02:03
Air Blank		0.000	02:05
Subject Sample	# 2	0.286	02:06
Breath Volume		2.003	02:06
Air Blank		0.000	02:07
Control Test		0.078	02:07
Air Blank		0.000	02:08
Diagnostics Check		OK	02:08

Data Download Date/Time: 11/21/2019 12:26

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETZ,FRANK,L
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: BETZ,FRANK
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 07:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:58

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 11/08/2019 08:24:45

Diagnostics Check		OK	08:18
Air Blank		0.000	08:19
Control Test		0.078	08:19
Air Blank		0.000	08:19
Subject Sample	# 1	0.142	08:20
Breath Volume		1.925	08:20
Air Blank		0.000	08:20
Air Blank		0.000	08:22
Subject Sample	# 2	0.143	08:23
Breath Volume		2.695	08:23
Air Blank		0.000	08:23
Control Test		0.077	08:24
Air Blank		0.000	08:24
Diagnostics Check		OK	08:24

Data Download Date/Time: 11/21/2019 12:26

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: CHLEBANOWSKI,RYAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: GRANT

DATE/TIME OF TEST: 11/09/2019 02:28:30

Diagnostics Check		OK	02:25	
Air Blank		0.000	02:26	
Control Test		0.080	02:26	
Air Blank		0.000	02:26	
Subject Sample	# 1		02:27	Subject Test Refused
Breath Volume		0.000	02:27	
Air Blank		0.000	02:27	
Control Test		0.079	02:27	
Air Blank		0.000	02:28	
Diagnostics Check		OK	02:28	

Data Download Date/Time: 11/21/2019 12:26

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANE,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: LANE,JONATHAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: ELLAHAM

DATE/TIME OF TEST: 11/10/2019 00:46:04

Diagnostics Check		OK	00:39
Air Blank		0.000	00:40
Control Test		0.080	00:40
Air Blank		0.000	00:41
Subject Sample	# 1	0.158	00:41
Breath Volume		2.871	00:41
Air Blank		0.000	00:42
Air Blank		0.000	00:43
Subject Sample	# 2	0.163	00:44
Breath Volume		2.921	00:44
Air Blank		0.000	00:45
Control Test		0.079	00:45
Air Blank		0.000	00:45
Diagnostics Check		OK	00:46

Data Download Date/Time: 11/21/2019 12:26

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUDSON,ANDREW,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: HUDSON,ANDREW
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: PFEIFFER

DATE/TIME OF TEST: 11/10/2019 23:39:49

Diagnostics Check		OK	23:26	
Air Blank		0.000	23:26	
Control Test		0.080	23:27	
Air Blank		0.000	23:27	
Subject Sample	# 1	0.194	23:28	
Breath Volume		1.265	23:28	
Air Blank		0.000	23:29	
Air Blank		0.000	23:30	
Subject Sample	# 2	0.149	23:33	Volume Not Met
Breath Volume		0.332	23:33	
Air Blank		0.000	23:34	
Air Blank		0.000	23:36	
Subject Sample	# 3		23:38	Subject Test Refused
Breath Volume		0.839	23:38	
Air Blank		0.000	23:38	
Control Test		0.077	23:39	
Air Blank		0.000	23:39	
Diagnostics Check		OK	23:39	

Data Download Date/Time: 11/21/2019 12:26

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEN,JOHN,M
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: GUIEB,ANTONIO
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 06:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:09

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 11/13/2019 07:47:26

Diagnostics Check		OK	07:41
Air Blank		0.000	07:41
Control Test		0.081	07:42
Air Blank		0.000	07:42
Subject Sample	# 1	0.000	07:43
Breath Volume		3.675	07:43
Air Blank		0.000	07:43
Air Blank		0.000	07:45
Subject Sample	# 2	0.000	07:46
Breath Volume		3.253	07:46
Air Blank		0.000	07:46
Control Test		0.080	07:46
Air Blank		0.000	07:47
Diagnostics Check		OK	07:47

Data Download Date/Time: 11/21/2019 12:26

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TELLIER,JEREMEY,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: DEHANAS,BRETT
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 11/14/2019 01:25:46

Diagnostics Check		OK	01:18
Air Blank		0.000	01:19
Control Test		0.080	01:19
Air Blank		0.000	01:19
Subject Sample	# 1	0.194	01:20
Breath Volume		1.187	01:20
Air Blank		0.000	01:21
Air Blank		0.000	01:22
Subject Sample	# 2	0.206	01:24
Breath Volume		1.773	01:24
Air Blank		0.000	01:24
Control Test		0.078	01:25
Air Blank		0.000	01:25
Diagnostics Check		OK	01:25

Data Download Date/Time: 11/21/2019 12:26

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARGUELLO,VICTOR,S
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: DEHANAS,BRETT
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 21:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:23

SUBJECT'S LAST NAME: SHARP

DATE/TIME OF TEST: 11/14/2019 22:46:47

Diagnostics Check		OK	22:43	
Air Blank		0.000	22:44	
Control Test		0.079	22:44	
Air Blank		0.000	22:45	
Subject Sample	# 1		22:45	Subject Test Refused
Breath Volume		0.000	22:45	
Air Blank		0.000	22:45	
Control Test		0.079	22:46	
Air Blank		0.000	22:46	
Diagnostics Check		OK	22:46	

Data Download Date/Time: 11/21/2019 12:26

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: CHLEBANOWSKI,RYAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:29

SUBJECT'S LAST NAME: BUFFINGTON

DATE/TIME OF TEST: 11/16/2019 00:31:07

Diagnostics Check	OK	00:30	
Air Blank		00:30	RFI Detect
Air Blank	0.000	00:31	

Data Download Date/Time: 11/21/2019 12:26

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: CHLEBANOWSKI,RYAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: BUFFINGTON

DATE/TIME OF TEST: 11/16/2019 00:58:22

Diagnostics Check		OK	00:54	
Air Blank		0.000	00:55	
Control Test		0.081	00:55	
Air Blank		0.000	00:56	
Subject Sample	# 1		00:56	Subject Test Refused
Breath Volume		0.000	00:56	
Air Blank		0.000	00:57	
Control Test		0.080	00:57	
Air Blank		0.000	00:58	
Diagnostics Check		OK	00:58	

Data Download Date/Time: 11/21/2019 12:26

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TYLER,BRYAN,D
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: TYLER,BRYAN
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: RUSSELL

DATE/TIME OF TEST: 11/16/2019 01:09:16

Diagnostics Check		OK	01:03
Air Blank		0.000	01:03
Control Test		0.080	01:04
Air Blank		0.000	01:04
Subject Sample	# 1	0.000	01:05
Breath Volume		1.949	01:05
Air Blank		0.000	01:05
Air Blank		0.000	01:07
Subject Sample	# 2	0.000	01:07
Breath Volume		2.160	01:07
Air Blank		0.000	01:08
Control Test		0.080	01:08
Air Blank		0.000	01:09
Diagnostics Check		OK	01:09

Data Download Date/Time: 11/21/2019 12:26

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOGERLEY,NEIL,F
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: MOGERLEY,NEIL
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: COLE

DATE/TIME OF TEST: 11/19/2019 02:25:32

Diagnostics Check		OK	02:18
Air Blank		0.000	02:18
Control Test		0.081	02:19
Air Blank		0.000	02:19
Subject Sample	# 1	0.177	02:20
Breath Volume		1.832	02:20
Air Blank		0.000	02:20
Air Blank		0.000	02:22
Subject Sample	# 2	0.169	02:23
Breath Volume		1.101	02:23
Air Blank		0.000	02:24
Control Test		0.080	02:24
Air Blank		0.000	02:25
Diagnostics Check		OK	02:25

Data Download Date/Time: 11/21/2019 12:26

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANE,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: LANE,JONATHAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 11/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: DOHRING

DATE/TIME OF TEST: 11/22/2019 21:27:29

Diagnostics Check		OK	21:21
Air Blank		0.000	21:21
Control Test		0.079	21:22
Air Blank		0.000	21:22
Subject Sample	# 1	0.202	21:22
Breath Volume		1.617	21:22
Air Blank		0.000	21:23
Air Blank		0.000	21:25
Subject Sample	# 2	0.200	21:25
Breath Volume		1.515	21:25
Air Blank		0.000	21:26
Control Test		0.079	21:26
Air Blank		0.000	21:27
Diagnostics Check		OK	21:27

Data Download Date/Time: 12/18/2019 14:38

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: CHLEBANOWSKI,RYAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 11/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: MICKELSON

DATE/TIME OF TEST: 11/22/2019 22:25:23

Diagnostics Check		OK	22:18
Air Blank		0.000	22:19
Control Test		0.080	22:19
Air Blank		0.000	22:20
Subject Sample	# 1	0.161	22:20
Breath Volume		1.132	22:20
Air Blank		0.000	22:21
Air Blank		0.000	22:23
Subject Sample	# 2	0.159	22:23
Breath Volume		1.296	22:23
Air Blank		0.000	22:24
Control Test		0.079	22:24
Air Blank		0.000	22:25
Diagnostics Check		OK	22:25

Data Download Date/Time: 12/18/2019 14:38

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOGERLEY,NEIL,F
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: MOGERLEY,NEIL
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 11/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: MARCHER

DATE/TIME OF TEST: 11/24/2019 01:52:16

Diagnostics Check		OK	01:45
Air Blank		0.000	01:46
Control Test		0.080	01:46
Air Blank		0.000	01:47
Subject Sample	# 1	0.184	01:47
Breath Volume		1.742	01:47
Air Blank		0.000	01:48
Air Blank		0.000	01:49
Subject Sample	# 2	0.197	01:50
Breath Volume		4.351	01:50
Air Blank		0.000	01:51
Control Test		0.079	01:51
Air Blank		0.000	01:52
Diagnostics Check		OK	01:52

Data Download Date/Time: 12/18/2019 14:38

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TELLIER,JEREMEY,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: TELLIER,JEREMEY
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 11/21/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: M450067792633

DATE/TIME OF TEST: 11/24/2019 04:52:39

Diagnostics Check		OK	04:45
Air Blank		0.000	04:46
Control Test		0.080	04:46
Air Blank		0.000	04:47
Subject Sample	# 1	0.149	04:48
Breath Volume		1.875	04:48
Air Blank		0.000	04:48
Air Blank		0.000	04:50
Subject Sample	# 2	0.145	04:50
Breath Volume		1.433	04:50
Air Blank		0.000	04:51
Control Test		0.077	04:52
Air Blank		0.000	04:52
Diagnostics Check		OK	04:52

Data Download Date/Time: 12/18/2019 14:38

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARGUELLO,VICTOR,S
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: ARGUELLO,VICTOR
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 11/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 11/24/2019 05:02:45

Diagnostics Check		OK	04:55
Air Blank		0.000	04:56
Control Test		0.079	04:56
Air Blank		0.000	04:56
Subject Sample	# 1	0.170	04:57
Breath Volume		1.523	04:57
Air Blank		0.000	04:58
Air Blank		0.000	05:00
Subject Sample	# 2	0.184	05:01
Breath Volume		2.664	05:01
Air Blank		0.000	05:01
Control Test		0.079	05:02
Air Blank		0.000	05:02
Diagnostics Check		OK	05:02

Data Download Date/Time: 12/18/2019 14:38

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUDSON,ANDREW,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: HUDSON,ANDREW
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 11/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 04:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: MORRISSEY

DATE/TIME OF TEST: 11/24/2019 05:11:06

Diagnostics Check		OK	05:08	
Air Blank		0.000	05:08	
Control Test		0.079	05:09	
Air Blank		0.000	05:09	
Subject Sample	# 1		05:09	Subject Test Refused
Breath Volume		0.000	05:09	
Air Blank		0.000	05:10	
Control Test		0.080	05:10	
Air Blank		0.000	05:11	
Diagnostics Check		OK	05:11	

Data Download Date/Time: 12/18/2019 14:38

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUDSON,ANDREW,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: LOPEZ,JORGE
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 11/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 11/25/2019 00:18:17

Diagnostics Check		OK	00:15	
Air Blank		0.000	00:15	
Control Test		0.080	00:16	
Air Blank		0.000	00:16	
Subject Sample	# 1		00:16	Subject Test Refused
Breath Volume		0.000	00:16	
Air Blank		0.000	00:17	
Control Test		0.080	00:17	
Air Blank		0.000	00:18	
Diagnostics Check		OK	00:18	

Data Download Date/Time: 12/18/2019 14:38

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANE,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: LANE,JONATHAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 11/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:03

SUBJECT'S LAST NAME: NEAUBAUER

DATE/TIME OF TEST: 11/28/2019 21:38:19

Diagnostics Check		OK	21:24	
Air Blank		0.000	21:24	
Control Test		0.080	21:24	
Air Blank		0.000	21:25	
Subject Sample	# 1	0.000	21:28	Volume Not Met
Breath Volume		0.832	21:28	
Air Blank		0.000	21:28	
Air Blank		0.000	21:30	
Subject Sample	# 2	0.000	21:31	
Breath Volume		1.179	21:31	
Air Blank		0.000	21:31	
Air Blank		0.000	21:33	
Subject Sample	# 3	0.000	21:36	Volume Not Met
Breath Volume		0.769	21:36	
Air Blank		0.000	21:37	
Control Test		0.080	21:37	
Air Blank		0.000	21:38	
Diagnostics Check		OK	21:38	

Data Download Date/Time: 12/18/2019 14:38

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: CHLEBANOWSKI,RYAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 11/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: GUTIERREZ COLLAZO

DATE/TIME OF TEST: 11/30/2019 22:46:27

Diagnostics Check		OK	22:38
Air Blank		0.000	22:38
Control Test		0.080	22:39
Air Blank		0.000	22:39
Subject Sample	# 1	0.165	22:41
Breath Volume		1.820	22:41
Air Blank		0.000	22:42
Air Blank		0.000	22:44
Subject Sample	# 2	0.158	22:44
Breath Volume		2.710	22:44
Air Blank		0.000	22:45
Control Test		0.077	22:45
Air Blank		0.000	22:46
Diagnostics Check		OK	22:46

Data Download Date/Time: 12/18/2019 14:38

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HRADECKY,CORBIN,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HRADECKY,CORBIN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 10/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: LISCIANDRO

DATE/TIME OF TEST: 11/06/2019 19:42:01

Diagnostics Check		OK	19:35
Air Blank		0.000	19:36
Control Test		0.078	19:36
Air Blank		0.000	19:36
Subject Sample	# 1	0.138	19:37
Breath Volume		1.695	19:37
Air Blank		0.000	19:38
Air Blank		0.000	19:39
Subject Sample	# 2	0.131	19:40
Breath Volume		1.710	19:40
Air Blank		0.000	19:41
Control Test		0.079	19:41
Air Blank		0.000	19:41
Diagnostics Check		OK	19:42

Data Download Date/Time: 11/20/2019 11:20

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HRADECKY,CORBIN,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: KIEFER,RUSSEL
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 10/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: BUI
ARREST TIME: 20:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:04

SUBJECT'S LAST NAME: ZENS

DATE/TIME OF TEST: 11/07/2019 21:41:11

Diagnostics Check		OK	21:24	
Air Blank		0.000	21:25	
Control Test		0.079	21:25	
Air Blank		0.000	21:25	
Subject Sample	# 1	0.136	21:28	
Breath Volume		1.515	21:28	
Air Blank		0.000	21:28	
Air Blank		0.000	21:30	
Subject Sample	# 2	0.139	21:33	Volume Not Met
Breath Volume		0.085	21:33	
Air Blank		0.000	21:34	
Air Blank		0.000	21:36	
Subject Sample	# 3	0.150	21:39	Volume Not Met
Breath Volume		0.054	21:39	
Air Blank		0.000	21:40	
Control Test		0.076	21:40	
Air Blank		0.000	21:41	
Diagnostics Check		OK	21:41	

Data Download Date/Time: 11/20/2019 11:20

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HRADECKY,CORBIN,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HRADECKY,CORBIN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 10/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: BRIGNOLO

DATE/TIME OF TEST: 11/08/2019 00:28:16

Diagnostics Check		OK	00:22
Air Blank		0.000	00:22
Control Test		0.078	00:22
Air Blank		0.000	00:23
Subject Sample	# 1	0.219	00:23
Breath Volume		2.992	00:23
Air Blank		0.000	00:24
Air Blank		0.000	00:26
Subject Sample	# 2	0.200	00:26
Breath Volume		1.933	00:26
Air Blank		0.000	00:27
Control Test		0.077	00:27
Air Blank		0.000	00:28
Diagnostics Check		OK	00:28

Data Download Date/Time: 11/20/2019 11:20

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HRADECKY,CORBIN,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HRADECKY,CORBIN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 10/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: CANDELARIO VILLAMAN

DATE/TIME OF TEST: 11/10/2019 02:23:12

Diagnostics Check		OK	02:16
Air Blank		0.000	02:16
Control Test		0.079	02:17
Air Blank		0.000	02:17
Subject Sample	# 1	0.198	02:18
Breath Volume		3.664	02:18
Air Blank		0.000	02:19
Air Blank		0.000	02:21
Subject Sample	# 2	0.197	02:21
Breath Volume		2.035	02:21
Air Blank		0.000	02:22
Control Test		0.078	02:22
Air Blank		0.000	02:23
Diagnostics Check		OK	02:23

Data Download Date/Time: 11/20/2019 11:20

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: FAISON,MEGAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 10/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: MARTINO

DATE/TIME OF TEST: 11/17/2019 00:57:55

Diagnostics Check		OK	00:52	
Air Blank		0.000	00:53	
Control Test		0.080	00:53	
Air Blank		0.000	00:54	
Subject Sample	# 1		00:56	Subject Test Refused
Breath Volume		0.000	00:56	
Air Blank		0.000	00:56	
Control Test		0.080	00:57	
Air Blank		0.000	00:57	
Diagnostics Check		OK	00:57	

Data Download Date/Time: 11/20/2019 11:20

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN-REYES,JUAN,C
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: AGUANNO,CHRISTOPHER
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 10/02/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 14:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:18

SUBJECT'S LAST NAME: SANABRIA CRUZ

DATE/TIME OF TEST: 11/17/2019 15:41:04

Diagnostics Check		OK	15:38	
Air Blank		0.000	15:38	
Control Test		0.081	15:38	
Air Blank		0.000	15:39	
Subject Sample	# 1		15:39	Subject Test Refused
Breath Volume		0.000	15:39	
Air Blank		0.000	15:40	
Control Test		0.080	15:40	
Air Blank		0.000	15:40	
Diagnostics Check		OK	15:41	

Data Download Date/Time: 11/20/2019 11:20

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HRADECKY,CORBIN,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HRADECKY,CORBIN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 10/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: ROA AGUILAR

DATE/TIME OF TEST: 11/20/2019 03:09:51

Diagnostics Check		OK	02:57	
Air Blank		0.000	02:57	
Control Test		0.080	02:58	
Air Blank		0.000	02:58	
Subject Sample	# 1	0.142	03:01	Volume Not Met
Breath Volume		0.292	03:01	
Air Blank		0.000	03:02	
Air Blank		0.000	03:04	
Subject Sample	# 2	0.154	03:05	
Breath Volume		1.156	03:05	
Air Blank		0.000	03:05	
Air Blank		0.000	03:07	
Subject Sample	# 3	0.163	03:08	
Breath Volume		1.335	03:08	
Air Blank		0.000	03:08	
Control Test		0.079	03:09	
Air Blank		0.000	03:09	
Diagnostics Check		OK	03:09	

Data Download Date/Time: 11/20/2019 11:20

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEIRD,ANDREW,D
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: FIERRO,ALEX
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 11/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 09:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:06

SUBJECT'S LAST NAME: ESTES

DATE/TIME OF TEST: 11/25/2019 10:33:19

Diagnostics Check		OK	10:26
Air Blank		0.000	10:27
Control Test		0.078	10:27
Air Blank		0.000	10:27
Subject Sample	# 1	0.331	10:28
Breath Volume		2.089	10:28
Air Blank		0.000	10:29
Air Blank		0.000	10:31
Subject Sample	# 2	0.320	10:31
Breath Volume		1.160	10:31
Air Blank		0.000	10:32
Control Test		0.078	10:32
Air Blank		0.000	10:33
Diagnostics Check		OK	10:33

Data Download Date/Time: 12/05/2019 14:36

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTIERREZ DEL TORO,RUBEN,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: WHITE,BRANDON
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 11/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: CASTELLANOS

DATE/TIME OF TEST: 11/28/2019 01:27:58

Diagnostics Check		OK	01:25	
Air Blank		0.000	01:25	
Control Test		0.080	01:25	
Air Blank		0.000	01:26	
Subject Sample	# 1		01:26	Subject Test Refused
Breath Volume		0.000	01:26	
Air Blank		0.000	01:27	
Control Test		0.079	01:27	
Air Blank		0.000	01:27	
Diagnostics Check		OK	01:27	

Data Download Date/Time: 12/05/2019 14:36

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONES,SHAUN,G
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: LONES,SHAUN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 11/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 15:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:37

SUBJECT'S LAST NAME: FAZEKAS

DATE/TIME OF TEST: 11/28/2019 16:04:37

Diagnostics Check		OK	15:57
Air Blank		0.000	15:58
Control Test		0.080	15:58
Air Blank		0.000	15:59
Subject Sample	# 1	0.213	15:59
Breath Volume		3.882	15:59
Air Blank		0.000	16:00
Air Blank		0.000	16:02
Subject Sample	# 2	0.196	16:02
Breath Volume		3.058	16:02
Air Blank		0.000	16:03
Control Test		0.079	16:03
Air Blank		0.000	16:04
Diagnostics Check		OK	16:04

Data Download Date/Time: 12/05/2019 14:36

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONES,SHAUN,G
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: LONES,SHAUN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 11/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: CHURCH

DATE/TIME OF TEST: 11/30/2019 19:42:09

Diagnostics Check		OK	19:35
Air Blank		0.000	19:35
Control Test		0.080	19:36
Air Blank		0.000	19:36
Subject Sample	# 1	0.185	19:37
Breath Volume		1.750	19:37
Air Blank		0.000	19:38
Air Blank		0.000	19:40
Subject Sample	# 2	0.182	19:40
Breath Volume		1.328	19:40
Air Blank		0.000	19:41
Control Test		0.078	19:41
Air Blank		0.000	19:42
Diagnostics Check		OK	19:42

Data Download Date/Time: 12/05/2019 14:36

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOEDICKE,WILLIAM,R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: JOEDICKE,WILLIAM
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 10/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: ROYER

DATE/TIME OF TEST: 11/02/2019 00:21:00

Diagnostics Check		OK	00:14
Air Blank		0.000	00:15
Control Test		0.078	00:15
Air Blank		0.000	00:15
Subject Sample	# 1	0.131	00:16
Breath Volume		1.382	00:16
Air Blank		0.000	00:17
Air Blank		0.000	00:18
Subject Sample	# 2	0.138	00:19
Breath Volume		3.050	00:19
Air Blank		0.000	00:20
Control Test		0.077	00:20
Air Blank		0.000	00:20
Diagnostics Check		OK	00:21

Data Download Date/Time: 11/19/2019 12:59

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,R,
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: BATTS,C
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 10/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: FYFFE

DATE/TIME OF TEST: 11/03/2019 00:47:15

Diagnostics Check		OK	00:43	
Air Blank		0.000	00:43	
Control Test		0.078	00:44	
Air Blank		0.000	00:44	
Subject Sample	# 1		00:45	Subject Test Refused
Breath Volume		0.000	00:45	
Air Blank		0.000	00:46	
Control Test		0.078	00:46	
Air Blank		0.000	00:47	
Diagnostics Check		OK	00:47	

Data Download Date/Time: 11/19/2019 12:59

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMARCO,RICHARD,E
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: DEMARCO,RICHARD
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 10/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: UHLENHAKE II

DATE/TIME OF TEST: 11/03/2019 03:56:59

Diagnostics Check		OK	03:49
Air Blank		0.000	03:49
Control Test		0.078	03:50
Air Blank		0.000	03:50
Subject Sample	# 1	0.125	03:52
Breath Volume		1.710	03:52
Air Blank		0.000	03:53
Air Blank		0.000	03:54
Subject Sample	# 2	0.126	03:55
Breath Volume		1.511	03:55
Air Blank		0.000	03:56
Control Test		0.076	03:56
Air Blank		0.000	03:56
Diagnostics Check		OK	03:56

Data Download Date/Time: 11/19/2019 12:59

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAINES,ELDON,L
BREATH TEST OPERATOR AGENCY: Dunnellon PD
ARRESTING OFFICER: RAINES,ELDON
ARRESTING OFFICER AGENCY: Dunnellon PD
LAST AGENCY INSPECTION: 10/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: MCLOCHLIN

DATE/TIME OF TEST: 11/03/2019 04:20:44

Diagnostics Check		OK	04:11	
Air Blank		0.000	04:12	
Control Test		0.078	04:12	
Air Blank		0.000	04:13	
Subject Sample	# 1	0.194	04:13	
Breath Volume		2.339	04:13	
Air Blank		0.000	04:14	
Air Blank		0.000	04:16	
Subject Sample	# 2	0.176	04:19	Volume Not Met
Breath Volume		0.121	04:19	
Air Blank			04:20	Purge Fail
Air Blank		0.000	04:20	

Data Download Date/Time: 11/19/2019 12:59

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAINES,ELDON,L
BREATH TEST OPERATOR AGENCY: Dunnellon PD
ARRESTING OFFICER: RAINES,ELDON
ARRESTING OFFICER AGENCY: Dunnellon PD
LAST AGENCY INSPECTION: 10/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: MCLOCHLIN

DATE/TIME OF TEST: 11/03/2019 04:24:12

Diagnostics Check	OK	04:22	
Air Blank	0.000	04:23	
Control Test	0.074	04:23	Control Outside Tolerance
Air Blank	0.000	04:24	

Data Download Date/Time: 11/19/2019 12:59

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAINES,ELDON,L
BREATH TEST OPERATOR AGENCY: Dunnellon PD
ARRESTING OFFICER: RAINES,ELDON
ARRESTING OFFICER AGENCY: Dunnellon PD
LAST AGENCY INSPECTION: 10/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: MCLOCHLIN

DATE/TIME OF TEST: 11/03/2019 04:37:22

Diagnostics Check		OK	04:25	
Air Blank		0.000	04:26	
Control Test		0.076	04:26	
Air Blank		0.000	04:26	
Subject Sample	# 1	0.181	04:27	
Breath Volume		1.281	04:27	
Air Blank		0.000	04:28	
Air Blank		0.000	04:30	
Subject Sample	# 2	0.160	04:32	
Breath Volume		1.464	04:32	
Air Blank		0.000	04:33	
Air Blank		0.000	04:34	
Subject Sample	# 3	0.171	04:35	
Breath Volume		1.207	04:35	
Air Blank		0.000	04:36	
Control Test		0.074	04:36	Control Outside Tolerance
Air Blank		0.000	04:37	

Data Download Date/Time: 11/19/2019 12:59

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCQUAIG,JASMINE,T
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MANCHESTER,MILTON
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 10/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: PARRISH

DATE/TIME OF TEST: 11/04/2019 00:19:38

Diagnostics Check		OK	00:13
Air Blank		0.000	00:14
Control Test		0.078	00:14
Air Blank		0.000	00:15
Subject Sample	# 1	0.000	00:15
Breath Volume		1.937	00:15
Air Blank		0.000	00:15
Air Blank		0.000	00:17
Subject Sample	# 2	0.000	00:18
Breath Volume		1.628	00:18
Air Blank		0.000	00:18
Control Test		0.076	00:19
Air Blank		0.000	00:19
Diagnostics Check		OK	00:19

Data Download Date/Time: 11/19/2019 12:59

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,MARK,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: ROBERTS,JASON
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 10:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:20

SUBJECT'S LAST NAME: LARREGUI DENNIS

DATE/TIME OF TEST: 11/05/2019 12:46:53

Diagnostics Check		OK	12:40
Air Blank		0.000	12:40
Control Test		0.078	12:41
Air Blank		0.000	12:41
Subject Sample	# 1	0.000	12:42
Breath Volume		2.250	12:42
Air Blank		0.000	12:42
Air Blank		0.000	12:44
Subject Sample	# 2	0.000	12:45
Breath Volume		2.050	12:45
Air Blank		0.000	12:45
Control Test		0.078	12:46
Air Blank		0.000	12:46
Diagnostics Check		OK	12:46

Data Download Date/Time: 11/19/2019 12:59

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,CATHERINE,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: HALL,CATHERINE
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: ALBRIGHT

DATE/TIME OF TEST: 11/05/2019 20:16:53

Diagnostics Check		OK	20:09	
Air Blank		0.000	20:10	
Control Test		0.078	20:10	
Air Blank		0.000	20:11	
Subject Sample	# 1	0.157	20:11	
Breath Volume		2.542	20:11	
Air Blank		0.000	20:12	
Air Blank		0.000	20:14	
Subject Sample	# 2	0.142	20:15	
Breath Volume		2.066	20:15	
Air Blank		0.000	20:15	
Control Test		0.074	20:16	Control Outside Tolerance
Air Blank		0.000	20:16	

Data Download Date/Time: 11/19/2019 12:59

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,CATHERINE,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: HALL,CATHERINE
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: ALBRIGHT

DATE/TIME OF TEST: 11/05/2019 20:26:03

Diagnostics Check		OK	20:19	
Air Blank		0.000	20:20	
Control Test		0.077	20:20	
Air Blank		0.000	20:20	
Subject Sample	# 1	0.153	20:21	
Breath Volume		1.781	20:21	
Air Blank		0.000	20:22	
Air Blank		0.000	20:23	
Subject Sample	# 2	0.148	20:24	
Breath Volume		2.832	20:24	
Air Blank		0.000	20:25	
Control Test		0.067	20:25	Control Outside Tolerance
Air Blank		0.000	20:26	

Data Download Date/Time: 11/19/2019 12:59

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,CATHERINE,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: HALL,CATHERINE
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: ALBRIGHT

DATE/TIME OF TEST: 11/05/2019 20:37:36

Diagnostics Check		OK	20:30
Air Blank		0.000	20:31
Control Test		0.078	20:31
Air Blank		0.000	20:32
Subject Sample	# 1	0.157	20:32
Breath Volume		2.593	20:32
Air Blank		0.000	20:33
Air Blank		0.000	20:35
Subject Sample	# 2	0.157	20:35
Breath Volume		2.535	20:35
Air Blank		0.000	20:36
Control Test		0.078	20:36
Air Blank		0.000	20:37
Diagnostics Check		OK	20:37

Data Download Date/Time: 11/19/2019 12:59

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DETWEILER,BRETT,R
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: DETWEILER,BRETT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 11/06/2019 01:13:54

Diagnostics Check		OK	01:07
Air Blank		0.000	01:07
Control Test		0.077	01:08
Air Blank		0.000	01:08
Subject Sample	# 1	0.241	01:09
Breath Volume		1.203	01:09
Air Blank		0.000	01:10
Air Blank		0.000	01:11
Subject Sample	# 2	0.222	01:12
Breath Volume		2.281	01:12
Air Blank		0.000	01:12
Control Test		0.075	01:13
Air Blank		0.000	01:13
Diagnostics Check		OK	01:13

Data Download Date/Time: 11/19/2019 12:59

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYS,M,W
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MAYS,M
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 09:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:30

SUBJECT'S LAST NAME: GOWERS

DATE/TIME OF TEST: 11/06/2019 11:05:02

Diagnostics Check		OK	10:50	
Air Blank		0.000	10:51	
Control Test		0.078	10:51	
Air Blank		0.000	10:52	
Subject Sample	# 1		10:55	No Sample Provided
Breath Volume		0.000	10:55	
Air Blank		0.000	10:55	
Air Blank		0.000	10:57	
Subject Sample	# 2	0.113	11:00	
Breath Volume		1.230	11:00	
Air Blank		0.000	11:00	
Air Blank		0.000	11:02	
Subject Sample	# 3	0.130	11:03	
Breath Volume		1.445	11:03	
Air Blank		0.000	11:04	
Control Test		0.078	11:04	
Air Blank		0.000	11:04	
Diagnostics Check		OK	11:05	

Data Download Date/Time: 11/19/2019 12:59

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,CHRISTINA,T
BREATH TEST OPERATOR AGENCY: Belleview PD
ARRESTING OFFICER: ESQUIVEL,CHRISTINA
ARRESTING OFFICER AGENCY: Belleview PD
LAST AGENCY INSPECTION: 10/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: RICHARD

DATE/TIME OF TEST: 11/07/2019 02:29:26

Diagnostics Check		OK	02:22
Air Blank		0.000	02:23
Control Test		0.078	02:23
Air Blank		0.000	02:24
Subject Sample	# 1	0.183	02:24
Breath Volume		1.992	02:24
Air Blank		0.000	02:25
Air Blank		0.000	02:27
Subject Sample	# 2	0.183	02:27
Breath Volume		1.394	02:27
Air Blank		0.000	02:28
Control Test		0.078	02:28
Air Blank		0.000	02:29
Diagnostics Check		OK	02:29

Data Download Date/Time: 11/19/2019 12:59

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORCHARD,DUSTIN,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SABO,JESSICA
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:38

SUBJECT'S LAST NAME: LYNN

DATE/TIME OF TEST: 11/07/2019 20:05:41

Diagnostics Check		OK	19:58
Air Blank		0.000	19:58
Control Test		0.078	19:59
Air Blank		0.000	19:59
Subject Sample	# 1	0.000	20:00
Breath Volume		1.839	20:00
Air Blank		0.000	20:01
Air Blank		0.000	20:03
Subject Sample	# 2	0.000	20:04
Breath Volume		2.503	20:04
Air Blank		0.000	20:04
Control Test		0.079	20:05
Air Blank		0.000	20:05
Diagnostics Check		OK	20:05

Data Download Date/Time: 11/19/2019 12:59

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURGESS,JAMES,W
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: BURGESS,JAMES
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 10/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: FERLAND

DATE/TIME OF TEST: 11/09/2019 02:52:21

Diagnostics Check		OK	02:49	
Air Blank		0.000	02:49	
Control Test		0.078	02:50	
Air Blank		0.000	02:50	
Subject Sample	# 1		02:50	Subject Test Refused
Breath Volume		0.000	02:50	
Air Blank		0.000	02:51	
Control Test		0.078	02:51	
Air Blank		0.000	02:52	
Diagnostics Check		OK	02:52	

Data Download Date/Time: 11/19/2019 12:59

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BECCUE,JOSHUA,B
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: BECCUE,JOSHUA
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 10/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: MURRAY

DATE/TIME OF TEST: 11/10/2019 02:23:14

Diagnostics Check		OK	02:20	
Air Blank		0.000	02:20	
Control Test		0.078	02:21	
Air Blank		0.000	02:21	
Subject Sample	# 1		02:21	Subject Test Refused
Breath Volume		0.000	02:21	
Air Blank		0.000	02:22	
Control Test		0.078	02:22	
Air Blank		0.000	02:23	
Diagnostics Check		OK	02:23	

Data Download Date/Time: 11/19/2019 12:59

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GROSSO,DYLAN,J
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: SMITH,MICAH
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 10/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: WILMERING

DATE/TIME OF TEST: 11/10/2019 02:52:14

Diagnostics Check		OK	02:45
Air Blank		0.000	02:46
Control Test		0.079	02:46
Air Blank		0.000	02:46
Subject Sample	# 1	0.142	02:47
Breath Volume		1.687	02:47
Air Blank		0.000	02:48
Air Blank		0.000	02:50
Subject Sample	# 2	0.140	02:50
Breath Volume		1.464	02:50
Air Blank		0.000	02:51
Control Test		0.078	02:51
Air Blank		0.000	02:52
Diagnostics Check		OK	02:52

Data Download Date/Time: 11/19/2019 12:59

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOYER,ZACHARY,M
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: HAZEL,PHIL
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 10/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:51

SUBJECT'S LAST NAME: MCMULLEN

DATE/TIME OF TEST: 11/10/2019 17:21:43

Diagnostics Check		OK	17:15
Air Blank		0.000	17:15
Control Test		0.078	17:15
Air Blank		0.000	17:16
Subject Sample	# 1	0.272	17:16
Breath Volume		2.691	17:16
Air Blank		0.000	17:17
Air Blank		0.000	17:19
Subject Sample	# 2	0.256	17:19
Breath Volume		1.226	17:19
Air Blank		0.000	17:20
Control Test		0.076	17:21
Air Blank		0.000	17:21
Diagnostics Check		OK	17:21

Data Download Date/Time: 11/19/2019 12:59

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GROSSO,DYLAN,J
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: GROSSO,DYLAN
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 10/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:50

SUBJECT'S LAST NAME: DIXON

DATE/TIME OF TEST: 11/11/2019 06:18:17

Diagnostics Check		OK	06:11
Air Blank		0.000	06:12
Control Test		0.077	06:12
Air Blank		0.000	06:13
Subject Sample	# 1	0.189	06:13
Breath Volume		1.878	06:13
Air Blank		0.000	06:14
Air Blank		0.000	06:16
Subject Sample	# 2	0.193	06:16
Breath Volume		1.992	06:16
Air Blank		0.000	06:17
Control Test		0.078	06:17
Air Blank		0.000	06:18
Diagnostics Check		OK	06:18

Data Download Date/Time: 11/19/2019 12:59

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOOREHEAD,DAVID,L
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: GRIGGS,LEONARD
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 10/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 15:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:53

SUBJECT'S LAST NAME: LUGER

DATE/TIME OF TEST: 11/12/2019 16:27:15

Diagnostics Check		OK	16:14	
Air Blank		0.000	16:14	
Control Test		0.078	16:14	
Air Blank		0.000	16:15	
Subject Sample	# 1	0.106	16:17	
Breath Volume		1.335	16:17	
Air Blank		0.000	16:17	
Air Blank		0.000	16:19	
Subject Sample	# 2	0.093	16:22	Volume Not Met
Breath Volume		0.695	16:22	
Air Blank		0.000	16:23	
Air Blank		0.000	16:25	
Subject Sample	# 3	0.097	16:25	
Breath Volume		1.277	16:25	
Air Blank		0.000	16:26	
Control Test		0.075	16:26	
Air Blank		0.000	16:27	
Diagnostics Check		OK	16:27	

Data Download Date/Time: 11/19/2019 12:59

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORCHARD,DUSTIN,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: HILLIARD,JEFF
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 15:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: NEWMAN

DATE/TIME OF TEST: 11/13/2019 17:41:47

Diagnostics Check		OK	17:35
Air Blank		0.000	17:35
Control Test		0.078	17:36
Air Blank		0.000	17:36
Subject Sample	# 1	0.216	17:37
Breath Volume		2.121	17:37
Air Blank		0.000	17:37
Air Blank		0.000	17:39
Subject Sample	# 2	0.220	17:40
Breath Volume		2.171	17:40
Air Blank		0.000	17:40
Control Test		0.078	17:41
Air Blank		0.000	17:41
Diagnostics Check		OK	17:41

Data Download Date/Time: 11/19/2019 12:59

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWMAN,AUBREY,B
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: ROSSI,MICHAEL
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 10/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 10:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:38

SUBJECT'S LAST NAME: FISCHKELTA

DATE/TIME OF TEST: 11/14/2019 11:10:44

Diagnostics Check		OK	11:03
Air Blank		0.000	11:04
Control Test		0.078	11:04
Air Blank		0.000	11:04
Subject Sample	# 1	0.098	11:05
Breath Volume		3.140	11:05
Air Blank		0.000	11:06
Air Blank		0.000	11:08
Subject Sample	# 2	0.094	11:09
Breath Volume		3.386	11:09
Air Blank		0.000	11:09
Control Test		0.078	11:10
Air Blank		0.000	11:10
Diagnostics Check		OK	11:10

Data Download Date/Time: 11/19/2019 12:59

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCQUAIG,JASMINE,T
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: KELLY,RICHARD
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 10/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 11/16/2019 02:17:20

Diagnostics Check		OK	02:05	
Air Blank		0.000	02:05	
Control Test		0.079	02:06	
Air Blank		0.000	02:06	
Subject Sample	# 1		02:09	No Sample Provided
Breath Volume		0.000	02:09	
Air Blank		0.000	02:10	
Air Blank		0.000	02:12	
Subject Sample	# 2	0.116	02:12	
Breath Volume		1.937	02:12	
Air Blank		0.000	02:13	
Air Blank		0.000	02:15	
Subject Sample	# 3	0.111	02:15	
Breath Volume		2.019	02:15	
Air Blank		0.000	02:16	
Control Test		0.076	02:16	
Air Blank		0.000	02:17	
Diagnostics Check		OK	02:17	

Data Download Date/Time: 11/19/2019 12:59

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWENS,RYAN,D
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: DRAKE,DENISE
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 10/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 11/16/2019 03:18:20

Diagnostics Check		OK	03:11
Air Blank		0.000	03:12
Control Test		0.078	03:12
Air Blank		0.000	03:12
Subject Sample	# 1	0.208	03:13
Breath Volume		1.238	03:13
Air Blank		0.000	03:14
Air Blank		0.000	03:16
Subject Sample	# 2	0.202	03:16
Breath Volume		2.601	03:16
Air Blank		0.000	03:17
Control Test		0.077	03:17
Air Blank		0.000	03:18
Diagnostics Check		OK	03:18

Data Download Date/Time: 11/19/2019 12:59

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,CHRISTINA,T
BREATH TEST OPERATOR AGENCY: Belleview PD
ARRESTING OFFICER: ESQUIVEL,CHRISTINA
ARRESTING OFFICER AGENCY: Belleview PD
LAST AGENCY INSPECTION: 10/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: RIVERA PAGAN

DATE/TIME OF TEST: 11/16/2019 04:58:36

Diagnostics Check		OK	04:51
Air Blank		0.000	04:52
Control Test		0.078	04:52
Air Blank		0.000	04:53
Subject Sample	# 1	0.051	04:54
Breath Volume		2.031	04:54
Air Blank		0.000	04:54
Air Blank		0.000	04:56
Subject Sample	# 2	0.047	04:57
Breath Volume		2.207	04:57
Air Blank		0.000	04:57
Control Test		0.079	04:57
Air Blank		0.000	04:58
Diagnostics Check		OK	04:58

Data Download Date/Time: 11/19/2019 12:59

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEST,R,
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: BATTS,C
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 10/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 11/17/2019 00:28:15

Diagnostics Check		OK	00:19	
Air Blank		0.000	00:19	
Control Test		0.079	00:19	
Air Blank		0.000	00:20	
Subject Sample	# 1	0.191	00:20	
Breath Volume		2.148	00:20	
Air Blank		0.000	00:21	
Air Blank		0.000	00:23	
Subject Sample	# 2	0.162	00:23	
Breath Volume		1.468	00:23	
Air Blank		0.000	00:24	
Air Blank		0.000	00:26	
Subject Sample	# 3	0.199	00:27	
Breath Volume		4.070	00:27	
Air Blank			00:27	Purge Fail
Air Blank		0.000	00:28	

Data Download Date/Time: 11/19/2019 12:59

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARR,LUCAS,B
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: KARR,LUCAS
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 10/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: HAYSTON

DATE/TIME OF TEST: 11/18/2019 00:48:50

Diagnostics Check		OK	00:38	
Air Blank		0.000	00:38	
Control Test		0.079	00:38	
Air Blank		0.000	00:39	
Subject Sample	# 1	0.131	00:41	
Breath Volume		1.105	00:41	
Air Blank		0.000	00:42	
Air Blank		0.000	00:44	
Subject Sample	# 2	0.109	00:47	Volume Not Met
Breath Volume		0.921	00:47	
Air Blank			00:48	Purge Fail
Air Blank		0.000	00:48	

Data Download Date/Time: 11/19/2019 12:59

Printed: 14 February 2020

COBRA ©1997-2014 CMI, Inc.