

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LUJAN,ARTURO,J  
BREATH TEST OPERATOR AGENCY: Coral Gables PD  
ARRESTING OFFICER: GONZALEZ,JUSTIN  
ARRESTING OFFICER AGENCY: Coral Gables PD  
LAST AGENCY INSPECTION: 11/19/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080AI  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: GONZALEZ VITON

DATE/TIME OF TEST: 11/20/2019 23:39:45

---

Diagnostics Check	OK	23:36	
Air Blank	0.000	23:36	
Control Test	0.081	23:37	
Air Blank	0.000	23:37	
Subject Sample	# 1	23:38	Interferent Detect
Breath Volume	2.394	23:38	
Air Blank	0.000	23:38	
Control Test	0.048	23:39	Interferent Detect
Air Blank	0.000	23:39	

---

Data Download Date/Time: 12/26/2019 11:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LUJAN,ARTURO,J  
BREATH TEST OPERATOR AGENCY: Coral Gables PD  
ARRESTING OFFICER: GONZALEZ,JUSTIN  
ARRESTING OFFICER AGENCY: Coral Gables PD  
LAST AGENCY INSPECTION: 11/19/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080AI  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: GONZALEZ VITON

DATE/TIME OF TEST: 11/20/2019 23:48:12

---

Diagnostics Check	OK	23:41
Air Blank	0.000	23:42
Control Test	0.081	23:42
Air Blank	0.000	23:43
Subject Sample # 1	0.170	23:43
Breath Volume	3.433	23:43
Air Blank	0.000	23:44
Air Blank	0.000	23:46
Subject Sample # 2	0.170	23:46
Breath Volume	3.195	23:46
Air Blank	0.000	23:47
Control Test	0.081	23:47
Air Blank	0.000	23:48
Diagnostics Check	OK	23:48

---

Data Download Date/Time: 12/26/2019 11:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RIVERA QUINONES,IGNACIO,  
BREATH TEST OPERATOR AGENCY: Coral Gables PD  
ARRESTING OFFICER: RIVERA QUINONES,IGNACIO  
ARRESTING OFFICER AGENCY: Coral Gables PD  
LAST AGENCY INSPECTION: 11/19/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080AI  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 11/29/2019 22:41:17

---

Diagnostics Check	OK	22:38	
Air Blank	0.000	22:38	
Control Test	0.082	22:39	
Air Blank	0.000	22:39	
Subject Sample	# 1	22:39	Subject Test Refused
Breath Volume	0.000	22:39	
Air Blank	0.000	22:40	
Control Test	0.081	22:40	
Air Blank	0.000	22:41	
Diagnostics Check	OK	22:41	

---

Data Download Date/Time: 12/26/2019 11:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHADWELL,WENDY,B  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: CHADWELL,WENDY  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 10/30/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 09:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 09:53

SUBJECT'S LAST NAME: GUELFU

DATE/TIME OF TEST: 11/02/2019 10:25:12

---

Diagnostics Check	OK	10:17
Air Blank	0.000	10:17
Control Test	0.080	10:18
Air Blank	0.000	10:18
Subject Sample # 1	0.059	10:19
Breath Volume	1.101	10:19
Air Blank	0.000	10:19
Air Blank	0.000	10:21
Subject Sample # 2	0.060	10:23
Breath Volume	1.542	10:23
Air Blank	0.000	10:24
Control Test	0.084	10:24
Air Blank	0.000	10:25
Diagnostics Check	OK	10:25

---

Data Download Date/Time: 11/04/2019 07:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHADWELL,WENDY,B  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: CHADWELL,WENDY  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 10/30/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:11  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: FUENTES

DATE/TIME OF TEST: 11/03/2019 00:05:19

---

Diagnostics Check	OK	23:58
Air Blank	0.000	23:59
Control Test	0.080	23:59
Air Blank	0.000	23:59
Subject Sample # 1	0.160	00:00
Breath Volume	1.613	00:00
Air Blank	0.000	00:01
Air Blank	0.000	00:03
Subject Sample # 2	0.160	00:03
Breath Volume	1.648	00:03
Air Blank	0.000	00:04
Control Test	0.080	00:04
Air Blank	0.000	00:05
Diagnostics Check	OK	00:05

---

Data Download Date/Time: 11/04/2019 07:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

INSTRUMENT REGISTERED TO: CALHOUN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ARROYO,WILFREDO,  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: ARROYO,WILFREDO  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 11/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29017080A3  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 04:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: ONEAL

DATE/TIME OF TEST: 11/17/2019 05:05:16

---

Diagnostics Check	OK	04:59
Air Blank	0.000	04:59
Control Test	0.080	04:59
Air Blank	0.000	05:00
Subject Sample # 1	0.137	05:00
Breath Volume	3.093	05:00
Air Blank	0.000	05:01
Air Blank	0.000	05:03
Subject Sample # 2	0.140	05:03
Breath Volume	2.742	05:03
Air Blank	0.000	05:04
Control Test	0.079	05:04
Air Blank	0.000	05:05
Diagnostics Check	OK	05:05

---

Data Download Date/Time: 12/12/2019 01:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

INSTRUMENT REGISTERED TO: HOMESTEAD PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SINCORE,MICHAEL,A  
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD  
ARRESTING OFFICER: PEREZ,RAUL  
ARRESTING OFFICER AGENCY: HOMESTEAD PD  
LAST AGENCY INSPECTION: 11/22/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1114319  
DRY GAS EXPIRY DATE: 03/27/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 11/22/2019 22:46:57

---

Diagnostics Check	OK	22:40
Air Blank	0.000	22:41
Control Test	0.082	22:41
Air Blank	0.000	22:41
Subject Sample # 1	0.083	22:42
Breath Volume	1.519	22:42
Air Blank	0.000	22:43
Air Blank	0.000	22:44
Subject Sample # 2	0.079	22:45
Breath Volume	1.285	22:45
Air Blank	0.000	22:46
Control Test	0.082	22:46
Air Blank	0.000	22:46
Diagnostics Check	OK	22:46

---

Data Download Date/Time: 12/31/2019 16:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

INSTRUMENT REGISTERED TO: HOMESTEAD PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEREZ,RAUL,  
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD  
ARRESTING OFFICER: GILPIN,SHELISA  
ARRESTING OFFICER AGENCY: HOMESTEAD PD  
LAST AGENCY INSPECTION: 11/22/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1114319  
DRY GAS EXPIRY DATE: 03/27/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: GALAN MORENO

DATE/TIME OF TEST: 11/24/2019 01:42:19

---

Diagnostics Check	OK	01:33	
Air Blank	0.000	01:34	
Control Test	0.082	01:34	
Air Blank	0.000	01:35	
Subject Sample	# 1	01:38	No Sample Provided
Breath Volume	0.000	01:38	
Air Blank	0.000	01:38	
Air Blank	0.000	01:40	
Subject Sample	# 2	01:40	Subject Test Refused
Breath Volume	0.000	01:40	
Air Blank	0.000	01:41	
Control Test	0.082	01:41	
Air Blank	0.000	01:42	
Diagnostics Check	OK	01:42	

---

Data Download Date/Time: 12/31/2019 16:43



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

INSTRUMENT REGISTERED TO: HOMESTEAD PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GILPIN,SCHELISA,  
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD  
ARRESTING OFFICER: GILPIN,SCHELISA  
ARRESTING OFFICER AGENCY: HOMESTEAD PD  
LAST AGENCY INSPECTION: 11/22/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1114319  
DRY GAS EXPIRY DATE: 03/27/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: MARCOS

DATE/TIME OF TEST: 11/24/2019 02:57:51

---

Diagnostics Check		OK	02:50
Air Blank		0.000	02:50
Control Test		0.082	02:50
Air Blank		0.000	02:51
Subject Sample	# 1	0.256	02:52
Breath Volume		1.570	02:52
Air Blank		0.000	02:53
Air Blank		0.000	02:54
Subject Sample	# 2	0.250	02:56
Breath Volume		2.105	02:56
Air Blank		0.000	02:56
Control Test		0.082	02:57
Air Blank		0.000	02:57
Diagnostics Check		OK	02:57

---

Data Download Date/Time: 12/31/2019 16:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

INSTRUMENT REGISTERED TO: HOMESTEAD PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SINCORE,MICHAEL,A  
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD  
ARRESTING OFFICER: SINCORE,MICHAEL  
ARRESTING OFFICER AGENCY: HOMESTEAD PD  
LAST AGENCY INSPECTION: 11/22/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1114319  
DRY GAS EXPIRY DATE: 03/27/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: PEREZ GUZMAN

DATE/TIME OF TEST: 11/28/2019 22:37:12

---

Diagnostics Check	OK	22:34	
Air Blank	0.000	22:34	
Control Test	0.083	22:35	
Air Blank	0.000	22:35	
Subject Sample	# 1	22:35	Subject Test Refused
Breath Volume	0.000	22:35	
Air Blank	0.000	22:36	
Control Test	0.082	22:36	
Air Blank	0.000	22:37	
Diagnostics Check	OK	22:37	

---

Data Download Date/Time: 12/31/2019 16:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PAIGE,LEON,I  
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD  
ARRESTING OFFICER: PAIGE,LEON  
ARRESTING OFFICER AGENCY: MIAMI BEACH PD  
LAST AGENCY INSPECTION: 10/29/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814181  
DRY GAS EXPIRY DATE: 07/24/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 11/01/2019 03:57:49

---

Diagnostics Check	OK	03:53	
Air Blank	0.000	03:53	
Control Test	0.081	03:53	
Air Blank	0.000	03:54	
Subject Sample	# 1	03:56	Subject Test Refused
Breath Volume	0.000	03:56	
Air Blank	0.000	03:56	
Control Test	0.079	03:57	
Air Blank	0.000	03:57	
Diagnostics Check	OK	03:57	

---

Data Download Date/Time: 11/15/2019 11:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PAIGE,LEON,I  
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD  
ARRESTING OFFICER: PAIGE,LEON  
ARRESTING OFFICER AGENCY: MIAMI BEACH PD  
LAST AGENCY INSPECTION: 10/29/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814181  
DRY GAS EXPIRY DATE: 07/24/2022  
VIOLATION CODE: DUI  
ARREST TIME: 07:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:25

SUBJECT'S LAST NAME: KARMALI

DATE/TIME OF TEST: 11/01/2019 07:52:01

---

Diagnostics Check	OK	07:45
Air Blank	0.000	07:45
Control Test	0.081	07:46
Air Blank	0.000	07:46
Subject Sample # 1	0.180	07:47
Breath Volume	2.214	07:47
Air Blank	0.000	07:48
Air Blank	0.000	07:50
Subject Sample # 2	0.169	07:50
Breath Volume	1.945	07:50
Air Blank	0.000	07:51
Control Test	0.081	07:51
Air Blank	0.000	07:51
Diagnostics Check	OK	07:52

---

Data Download Date/Time: 11/15/2019 11:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CREWS,CURTIS,L  
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD  
ARRESTING OFFICER: PAIGE,LEON  
ARRESTING OFFICER AGENCY: MIAMI BEACH PD  
LAST AGENCY INSPECTION: 10/29/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1175968  
DRY GAS EXPIRY DATE: 07/30/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:39  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: PEREZ MIGUEL

DATE/TIME OF TEST: 11/02/2019 23:28:01

---

Diagnostics Check	OK	23:25	
Air Blank	0.000	23:25	
Control Test	0.081	23:25	
Air Blank	0.000	23:26	
Subject Sample	# 1	23:26	Subject Test Refused
Breath Volume	0.000	23:26	
Air Blank	0.000	23:27	
Control Test	0.080	23:27	
Air Blank	0.000	23:27	
Diagnostics Check	OK	23:28	

---

Data Download Date/Time: 11/15/2019 11:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PAIGE,LEON,I  
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD  
ARRESTING OFFICER: PAIGE,LEON  
ARRESTING OFFICER AGENCY: MIAMI BEACH PD  
LAST AGENCY INSPECTION: 10/29/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814181  
DRY GAS EXPIRY DATE: 07/24/2022  
VIOLATION CODE: DUI  
ARREST TIME: 05:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:56

SUBJECT'S LAST NAME: BURKLEY

DATE/TIME OF TEST: 11/04/2019 06:22:23

---

Diagnostics Check		OK	06:16
Air Blank		0.000	06:16
Control Test		0.082	06:17
Air Blank		0.000	06:17
Subject Sample	# 1	0.213	06:18
Breath Volume		1.750	06:18
Air Blank		0.000	06:18
Air Blank		0.000	06:20
Subject Sample	# 2	0.219	06:20
Breath Volume		3.863	06:20
Air Blank		0.000	06:21
Control Test		0.081	06:21
Air Blank		0.000	06:22
Diagnostics Check		OK	06:22

---

Data Download Date/Time: 11/15/2019 11:54

**Florida Department of Law Enforcement**

**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRIND,ABEL,D  
BREATH TEST OPERATOR AGENCY: NORTH BAY VILLAGE PD  
ARRESTING OFFICER: BRIND,ABEL  
ARRESTING OFFICER AGENCY: NORTH BAY VILLAGE PD  
LAST AGENCY INSPECTION: 10/29/2019  
Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814181  
DRY GAS EXPIRY DATE: 07/24/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:58  
SUBJECT'S LAST NAME: MULET  
DATE/TIME OF TEST: 11/08/2019 01:28:07

---

Diagnostics Check		OK	01:19
Air Blank		0.000	01:20
Control Test		0.082	01:20
Air Blank		0.000	01:20
Subject Sample	# 1	0.206	01:21
Breath Volume		1.234	01:21
Air Blank		0.000	01:21
Air Blank		0.000	01:23
Subject Sample	# 2	0.219	01:26
Breath Volume		1.675	01:26
Air Blank		0.000	01:27
Control Test		0.082	01:27
Air Blank		0.000	01:28
Diagnostics Check		OK	01:28

---

Data Download Date/Time: 11/15/2019 11:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MELENDEZ,DANIEL,  
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD  
ARRESTING OFFICER: MELENDEZ,DANIEL  
ARRESTING OFFICER AGENCY: MIAMI BEACH PD  
LAST AGENCY INSPECTION: 10/29/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814181  
DRY GAS EXPIRY DATE: 07/24/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 11/09/2019 00:38:37

---

Diagnostics Check	OK	00:35	
Air Blank	0.000	00:36	
Control Test	0.082	00:36	
Air Blank	0.000	00:37	
Subject Sample	# 1	00:37	Subject Test Refused
Breath Volume	0.000	00:37	
Air Blank	0.000	00:37	
Control Test	0.082	00:38	
Air Blank	0.000	00:38	
Diagnostics Check	OK	00:38	

---

Data Download Date/Time: 11/15/2019 11:54



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MELENDEZ,DANIEL,  
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD  
ARRESTING OFFICER: MELENDEZ,DANIEL  
ARRESTING OFFICER AGENCY: MIAMI BEACH PD  
LAST AGENCY INSPECTION: 10/29/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814181  
DRY GAS EXPIRY DATE: 07/24/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: ROJAS

DATE/TIME OF TEST: 11/10/2019 03:13:01

---

Diagnostics Check	OK	03:10	
Air Blank	0.000	03:10	
Control Test	0.082	03:11	
Air Blank	0.000	03:11	
Subject Sample	# 1	03:11	Subject Test Refused
Breath Volume	0.000	03:11	
Air Blank	0.000	03:12	
Control Test	0.081	03:12	
Air Blank	0.000	03:12	
Diagnostics Check	OK	03:13	

---

Data Download Date/Time: 11/15/2019 11:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MELENDEZ,DANIEL,  
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD  
ARRESTING OFFICER: BROWN,STEPHANIE  
ARRESTING OFFICER AGENCY: MIAMI BEACH PD  
LAST AGENCY INSPECTION: 10/29/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814181  
DRY GAS EXPIRY DATE: 07/24/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: LANDESMAN

DATE/TIME OF TEST: 11/10/2019 03:34:27

---

Diagnostics Check		OK	03:26
Air Blank		0.000	03:26
Control Test		0.082	03:26
Air Blank		0.000	03:27
Subject Sample	# 1	0.188	03:27
Breath Volume		1.433	03:27
Air Blank		0.000	03:28
Air Blank		0.000	03:30
Subject Sample	# 2	0.191	03:32
Breath Volume		1.839	03:32
Air Blank		0.000	03:33
Control Test		0.081	03:33
Air Blank		0.000	03:34
Diagnostics Check		OK	03:34

---

Data Download Date/Time: 11/15/2019 11:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUMPHRY,KERRY,R  
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD  
ARRESTING OFFICER: LERGIER,L  
ARRESTING OFFICER AGENCY: MIAMI BEACH PD  
LAST AGENCY INSPECTION: 10/29/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814181  
DRY GAS EXPIRY DATE: 07/24/2020  
VIOLATION CODE: DUI  
ARREST TIME: 06:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:20

SUBJECT'S LAST NAME: COHEN

DATE/TIME OF TEST: 11/12/2019 07:52:44

---

Diagnostics Check	OK	07:46
Air Blank	0.000	07:46
Control Test	0.082	07:47
Air Blank	0.000	07:47
Subject Sample # 1	0.190	07:48
Breath Volume	1.847	07:48
Air Blank	0.000	07:48
Air Blank	0.000	07:50
Subject Sample # 2	0.185	07:51
Breath Volume	1.410	07:51
Air Blank	0.000	07:51
Control Test	0.081	07:52
Air Blank	0.000	07:52
Diagnostics Check	OK	07:52

---

Data Download Date/Time: 11/15/2019 11:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MELENDEZ,DANIEL,  
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD  
ARRESTING OFFICER: PAIGE,LEON  
ARRESTING OFFICER AGENCY: MIAMI BEACH PD  
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814181  
DRY GAS EXPIRY DATE: 07/24/2020  
VIOLATION CODE: DUI  
ARREST TIME: 04:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:36

SUBJECT'S LAST NAME: TRUJILLO

DATE/TIME OF TEST: 11/16/2019 05:00:14

---

Diagnostics Check	OK	04:57	
Air Blank	0.000	04:57	
Control Test	0.079	04:58	
Air Blank	0.000	04:58	
Subject Sample	# 1	04:58	Subject Test Refused
Breath Volume	0.000	04:58	
Air Blank	0.000	04:59	
Control Test	0.080	04:59	
Air Blank	0.000	05:00	
Diagnostics Check	OK	05:00	

---

Data Download Date/Time: 12/26/2019 14:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CASTILLO, RICARDO, A  
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD  
ARRESTING OFFICER: CASTILLO, RICARDO  
ARRESTING OFFICER AGENCY: MIAMI BEACH PD  
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1175968  
DRY GAS EXPIRY DATE: 07/30/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:52  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:11

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 11/16/2019 22:35:04

---

Diagnostics Check	OK	22:32	
Air Blank	0.000	22:32	
Control Test	0.081	22:33	
Air Blank	0.000	22:33	
Subject Sample	# 1	22:33	Subject Test Refused
Breath Volume	0.000	22:33	
Air Blank	0.000	22:34	
Control Test	0.082	22:34	
Air Blank	0.000	22:34	
Diagnostics Check	OK	22:35	

---

Data Download Date/Time: 12/26/2019 14:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FERNANDEZ,GIRANNY,  
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD  
ARRESTING OFFICER: FERNANDEZ,GIRANNY  
ARRESTING OFFICER AGENCY: MIAMI BEACH PD  
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814181  
DRY GAS EXPIRY DATE: 07/24/2020  
VIOLATION CODE: DUI  
ARREST TIME: 18:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:48

SUBJECT'S LAST NAME: VASQUEZ

DATE/TIME OF TEST: 11/17/2019 19:15:30

---

Diagnostics Check	OK	19:12	
Air Blank	0.000	19:13	
Control Test	0.082	19:13	
Air Blank	0.000	19:13	
Subject Sample	# 1	19:14	Subject Test Refused
Breath Volume	0.000	19:14	
Air Blank	0.000	19:14	
Control Test	0.082	19:14	
Air Blank	0.000	19:15	
Diagnostics Check	OK	19:15	

---

Data Download Date/Time: 12/26/2019 14:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MELENDEZ,DANIEL,  
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD  
ARRESTING OFFICER: MELENDEZ,DANIEL  
ARRESTING OFFICER AGENCY: MIAMI BEACH PD  
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814181  
DRY GAS EXPIRY DATE: 07/24/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: PACHECO PEREZ

DATE/TIME OF TEST: 11/18/2019 02:13:33

---

Diagnostics Check	OK	02:07
Air Blank	0.000	02:07
Control Test	0.082	02:08
Air Blank	0.000	02:08
Subject Sample # 1	0.115	02:09
Breath Volume	3.511	02:09
Air Blank	0.000	02:09
Air Blank	0.000	02:11
Subject Sample # 2	0.115	02:12
Breath Volume	3.234	02:12
Air Blank	0.000	02:12
Control Test	0.081	02:12
Air Blank	0.000	02:13
Diagnostics Check	OK	02:13

---

Data Download Date/Time: 12/26/2019 14:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PAIGE,LEON,I  
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD  
ARRESTING OFFICER: PAIGE,LEON  
ARRESTING OFFICER AGENCY: MIAMI BEACH PD  
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814181  
DRY GAS EXPIRY DATE: 07/30/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: JAMHOUR

DATE/TIME OF TEST: 11/22/2019 00:34:28

---

Diagnostics Check		OK	00:26	
Air Blank		0.000	00:26	
Control Test		0.081	00:26	
Air Blank		0.000	00:27	
Subject Sample	# 1	0.154	00:30	Volume Not Met
Breath Volume		0.167	00:30	
Air Blank		0.000	00:31	
Air Blank		0.000	00:32	
Subject Sample	# 2		00:33	Subject Test Refused
Breath Volume		0.000	00:33	
Air Blank		0.000	00:33	
Control Test		0.079	00:33	
Air Blank		0.000	00:34	
Diagnostics Check		OK	00:34	

---

Data Download Date/Time: 12/26/2019 14:34



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PAIGE,LEON,I  
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD  
ARRESTING OFFICER: PAIGE,LEON  
ARRESTING OFFICER AGENCY: MIAMI BEACH PD  
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814181  
DRY GAS EXPIRY DATE: 07/30/2022  
VIOLATION CODE: DUI  
ARREST TIME: 06:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:45

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 11/23/2019 07:11:32

---

Diagnostics Check	OK	07:05
Air Blank	0.000	07:05
Control Test	0.080	07:06
Air Blank	0.000	07:06
Subject Sample # 1	0.195	07:07
Breath Volume	2.417	07:07
Air Blank	0.000	07:07
Air Blank	0.000	07:09
Subject Sample # 2	0.196	07:10
Breath Volume	3.261	07:10
Air Blank	0.000	07:10
Control Test	0.081	07:10
Air Blank	0.000	07:11
Diagnostics Check	OK	07:11

---

Data Download Date/Time: 12/26/2019 14:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SOSA, JONATHAN,  
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD  
ARRESTING OFFICER: CHEVALIER, ISMAEL  
ARRESTING OFFICER AGENCY: NORTH BAY VILLAGE PD  
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1175968  
DRY GAS EXPIRY DATE: 07/30/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:02

SUBJECT'S LAST NAME: ALVARADO ALVARADO

DATE/TIME OF TEST: 11/23/2019 21:51:27

---

Diagnostics Check	OK	21:38	
Air Blank	0.000	21:38	
Control Test	0.081	21:38	
Air Blank	0.000	21:39	
Subject Sample	# 1	21:42	No Sample Provided
Breath Volume	0.000	21:42	
Air Blank	0.000	21:43	
Air Blank	0.000	21:45	
Subject Sample	# 2	0.187	21:46
Breath Volume	1.460	21:46	
Air Blank	0.000	21:47	
Air Blank	0.000	21:49	
Subject Sample	# 3	0.192	21:49
Breath Volume	1.679	21:49	
Air Blank	0.000	21:50	
Control Test	0.082	21:50	
Air Blank	0.000	21:51	
Diagnostics Check	OK	21:51	

---

Data Download Date/Time: 12/26/2019 14:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,  
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD  
ARRESTING OFFICER: DOMINGUEZ,ERIC  
ARRESTING OFFICER AGENCY: MIAMI BEACH PD  
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814181  
DRY GAS EXPIRY DATE: 07/24/2020  
VIOLATION CODE: DUI  
ARREST TIME: 07:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:20

SUBJECT'S LAST NAME: RAKAH

DATE/TIME OF TEST: 11/28/2019 07:48:18

---

Diagnostics Check	OK	07:41
Air Blank	0.000	07:41
Control Test	0.081	07:41
Air Blank	0.000	07:42
Subject Sample # 1	0.137	07:43
Breath Volume	2.113	07:43
Air Blank	0.000	07:44
Air Blank	0.000	07:46
Subject Sample # 2	0.136	07:46
Breath Volume	1.976	07:46
Air Blank	0.000	07:47
Control Test	0.080	07:47
Air Blank	0.000	07:48
Diagnostics Check	OK	07:48

---

Data Download Date/Time: 12/26/2019 14:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AGINS,LYLE,P  
BREATH TEST OPERATOR AGENCY: Sunny Isles Beach PD  
ARRESTING OFFICER: DELGADO,M  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: ARCHIBALD

DATE/TIME OF TEST: 11/17/2019 23:19:58

---

Diagnostics Check	OK	23:13
Air Blank	0.000	23:14
Control Test	0.080	23:14
Air Blank	0.000	23:14
Subject Sample # 1	0.136	23:15
Breath Volume	1.292	23:15
Air Blank	0.000	23:15
Air Blank	0.000	23:17
Subject Sample # 2	0.139	23:18
Breath Volume	2.289	23:18
Air Blank	0.000	23:19
Control Test	0.079	23:19
Air Blank	0.000	23:19
Diagnostics Check	OK	23:19

---

Data Download Date/Time: 12/26/2019 12:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AGINS,LYLE,P  
BREATH TEST OPERATOR AGENCY: Sunny Isles Beach PD  
ARRESTING OFFICER: AGINS,LYLE  
ARRESTING OFFICER AGENCY: Sunny Isles Beach PD  
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: STATELMAN

DATE/TIME OF TEST: 11/24/2019 03:41:58

---

Diagnostics Check	OK	03:38	
Air Blank	0.000	03:38	
Control Test	0.078	03:38	
Air Blank	0.000	03:39	
Subject Sample	# 1	03:40	Subject Test Refused
Breath Volume	0.000	03:40	
Air Blank	0.000	03:41	
Control Test	0.080	03:41	
Air Blank	0.000	03:41	
Diagnostics Check	OK	03:41	

---

Data Download Date/Time: 12/26/2019 12:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ADAMS,STEVEN,L  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: ADAMS,STEVEN  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: GUTIERREZ

DATE/TIME OF TEST: 11/20/2019 04:05:48

---

Diagnostics Check	OK	04:02	
Air Blank	0.000	04:02	
Control Test	0.081	04:03	
Air Blank	0.000	04:03	
Subject Sample	# 1	04:04	Subject Test Refused
Breath Volume	0.000	04:04	
Air Blank	0.000	04:04	
Control Test	0.080	04:05	
Air Blank	0.000	04:05	
Diagnostics Check	OK	04:05	

---

Data Download Date/Time: 12/26/2019 12:55

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SUAREZ,CARLOS,A  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: SUAREZ,CARLOS  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 10/30/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:02

SUBJECT'S LAST NAME: RIVERA-MONCALEANO

DATE/TIME OF TEST: 11/03/2019 05:25:02

---

Diagnostics Check	OK	05:22	
Air Blank	0.000	05:22	
Control Test	0.079	05:22	
Air Blank	0.000	05:23	
Subject Sample	# 1	05:23	Subject Test Refused
Breath Volume	0.000	05:23	
Air Blank	0.000	05:24	
Control Test	0.080	05:24	
Air Blank	0.000	05:24	
Diagnostics Check	OK	05:25	

---

Data Download Date/Time: 11/15/2019 12:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WINLAND,JUSTIN,D  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: SELENT,SCOTT  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 10/30/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 13:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 12:30

SUBJECT'S LAST NAME: GALARZA

DATE/TIME OF TEST: 11/05/2019 13:13:03

---

Diagnostics Check	OK	13:12	
Air Blank	0.000	13:12	
Control Test	0.080	13:12	
Air Blank		13:12	RFI Detect
Air Blank		13:13	RFI Detect

---

Data Download Date/Time: 11/15/2019 12:19



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WINLAND,JUSTIN,D  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: SELENT,SCOTT  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 10/30/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 13:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:10

SUBJECT'S LAST NAME: GALARZA

DATE/TIME OF TEST: 11/05/2019 13:33:48

---

Diagnostics Check	OK	13:30	
Air Blank	0.000	13:31	
Control Test	0.081	13:31	
Air Blank	0.000	13:32	
Subject Sample	# 1	13:32	Subject Test Refused
Breath Volume	0.000	13:32	
Air Blank	0.000	13:32	
Control Test	0.080	13:33	
Air Blank	0.000	13:33	
Diagnostics Check	OK	13:33	

---

Data Download Date/Time: 11/15/2019 12:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GUTIERREZ,KRISTINA,L  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: QUINTANA,EDWARD  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 10/30/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 06:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:55

SUBJECT'S LAST NAME: ALVAREZ IXCOT

DATE/TIME OF TEST: 11/09/2019 07:22:18

---

Diagnostics Check		OK	07:15
Air Blank		0.000	07:15
Control Test		0.080	07:16
Air Blank		0.000	07:16
Subject Sample	# 1	0.241	07:17
Breath Volume		2.960	07:17
Air Blank		0.000	07:18
Air Blank		0.000	07:20
Subject Sample	# 2	0.230	07:20
Breath Volume		2.230	07:20
Air Blank		0.000	07:21
Control Test		0.079	07:21
Air Blank		0.000	07:22
Diagnostics Check		OK	07:22

---

Data Download Date/Time: 11/15/2019 12:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SUAREZ,CARLOS',A  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: SUAREZ,CARLOS  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 10/30/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: DELGADO-LOPEZ

DATE/TIME OF TEST: 11/11/2019 00:50:15

---

Diagnostics Check	OK	00:41	
Air Blank	0.000	00:42	
Control Test	0.080	00:42	
Air Blank	0.000	00:42	
Subject Sample	# 1 0.147	00:45	
Breath Volume	1.109	00:45	
Air Blank	0.000	00:45	
Air Blank	0.000	00:47	
Subject Sample	# 2	00:48	RFI Detect
Breath Volume	0.355	00:48	
Air Blank	0.000	00:49	
Control Test	0.080	00:49	
Air Blank	0.000	00:50	
Diagnostics Check	OK	00:50	

---

Data Download Date/Time: 11/15/2019 12:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SUAREZ,CARLOS,A  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: SUAREZ,CARLOS  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 10/30/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: DELGADO-LOPEZ

DATE/TIME OF TEST: 11/11/2019 00:59:54

---

Diagnostics Check	OK	00:52
Air Blank	0.000	00:52
Control Test	0.079	00:53
Air Blank	0.000	00:53
Subject Sample # 1	0.153	00:54
Breath Volume	1.183	00:54
Air Blank	0.000	00:55
Air Blank	0.000	00:57
Subject Sample # 2	0.153	00:58
Breath Volume	2.718	00:58
Air Blank	0.000	00:58
Control Test	0.079	00:59
Air Blank	0.000	00:59
Diagnostics Check	OK	00:59

---

Data Download Date/Time: 11/15/2019 12:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SUAREZ,CARLOS,A  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: DE LA OSSA,SUSANA  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: UBEDA

DATE/TIME OF TEST: 11/18/2019 00:26:45

---

Diagnostics Check	OK	00:20
Air Blank	0.000	00:20
Control Test	0.081	00:20
Air Blank	0.000	00:21
Subject Sample # 1	0.228	00:22
Breath Volume	2.851	00:22
Air Blank	0.000	00:22
Air Blank	0.000	00:24
Subject Sample # 2	0.222	00:25
Breath Volume	2.492	00:25
Air Blank	0.000	00:25
Control Test	0.080	00:26
Air Blank	0.000	00:26
Diagnostics Check	OK	00:26

---

Data Download Date/Time: 12/26/2019 12:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FLORIAN,ROLANDO,I  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: FLORIAN,ROLANDO  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 17:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:21

SUBJECT'S LAST NAME: VELAZQUEZ

DATE/TIME OF TEST: 11/29/2019 18:11:36

---

Diagnostics Check	OK	18:05
Air Blank	0.000	18:05
Control Test	0.080	18:05
Air Blank	0.000	18:06
Subject Sample # 1	0.283	18:07
Breath Volume	3.351	18:07
Air Blank	0.000	18:07
Air Blank	0.000	18:09
Subject Sample # 2	0.288	18:09
Breath Volume	3.203	18:09
Air Blank	0.000	18:10
Control Test	0.079	18:11
Air Blank	0.000	18:11
Diagnostics Check	OK	18:11

---

Data Download Date/Time: 12/26/2019 12:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SUAREZ,CARLOS,A  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: TORUNO,KARYN  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: FUENTES

DATE/TIME OF TEST: 11/30/2019 22:37:09

---

Diagnostics Check	OK	22:30
Air Blank	0.000	22:31
Control Test	0.080	22:31
Air Blank	0.000	22:32
Subject Sample # 1	0.172	22:32
Breath Volume	1.933	22:32
Air Blank	0.000	22:33
Air Blank	0.000	22:35
Subject Sample # 2	0.167	22:35
Breath Volume	1.910	22:35
Air Blank	0.000	22:36
Control Test	0.080	22:36
Air Blank	0.000	22:37
Diagnostics Check	OK	22:37

---

Data Download Date/Time: 12/26/2019 12:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PHILPOT,C,  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: PHILPOT,C  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34917080A4  
DRY GAS EXPIRY DATE: 02/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:32

SUBJECT'S LAST NAME: PETERSON

DATE/TIME OF TEST: 11/26/2019 21:58:53

---

Diagnostics Check	OK	21:52
Air Blank	0.000	21:52
Control Test	0.078	21:53
Air Blank	0.000	21:53
Subject Sample # 1	0.258	21:54
Breath Volume	1.882	21:54
Air Blank	0.000	21:55
Air Blank	0.000	21:56
Subject Sample # 2	0.258	21:57
Breath Volume	2.179	21:57
Air Blank	0.000	21:57
Control Test	0.079	21:58
Air Blank	0.000	21:58
Diagnostics Check	OK	21:58

---

Data Download Date/Time: 12/26/2019 12:34



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAYES,BRITTNEY,B  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: HAYES,BRITTNEY  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34917080A4  
DRY GAS EXPIRY DATE: 02/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: RESENDE ROMERO

DATE/TIME OF TEST: 11/28/2019 03:05:13

---

Diagnostics Check	OK	03:00	
Air Blank	0.000	03:01	
Control Test	0.078	03:01	
Air Blank	0.000	03:02	
Subject Sample	# 1	03:04	Sequence Aborted
Breath Volume	0.000	03:04	
Air Blank	0.000	03:05	

---

Data Download Date/Time: 12/26/2019 12:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBLEDO MARQUEZ,MIGUEL,A  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: ONEAL,CHARLES  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 07:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:30

SUBJECT'S LAST NAME: PERTUZ DIAZ

DATE/TIME OF TEST: 11/16/2019 08:04:33

---

Diagnostics Check	OK	07:58
Air Blank	0.000	07:58
Control Test	0.080	07:59
Air Blank	0.000	07:59
Subject Sample # 1	0.185	07:59
Breath Volume	2.550	07:59
Air Blank	0.000	08:00
Air Blank	0.000	08:02
Subject Sample # 2	0.194	08:02
Breath Volume	3.285	08:02
Air Blank	0.000	08:03
Control Test	0.080	08:03
Air Blank	0.000	08:04
Diagnostics Check	OK	08:04

---

Data Download Date/Time: 12/26/2019 15:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HELFNER,ISAAC,S  
BREATH TEST OPERATOR AGENCY: Aventura PD  
ARRESTING OFFICER: HELFNER,ISAAC  
ARRESTING OFFICER AGENCY: Aventura PD  
LAST AGENCY INSPECTION: 10/28/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34917080A4  
DRY GAS EXPIRY DATE: 02/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 14:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 14:17

SUBJECT'S LAST NAME: ANTOSCHENKO

DATE/TIME OF TEST: 11/21/2019 15:27:15

---

Diagnostics Check	OK	15:16	
Air Blank	0.000	15:17	
Control Test	0.078	15:17	
Air Blank	0.000	15:18	
Subject Sample	# 1 0.183	15:21	Volume Not Met
Breath Volume	0.285	15:21	
Air Blank	0.000	15:21	
Air Blank	0.000	15:23	
Subject Sample	# 2	15:25	Subject Test Refused
Breath Volume	0.179	15:25	
Air Blank	0.000	15:26	
Control Test	0.075	15:26	
Air Blank	0.000	15:27	
Diagnostics Check	OK	15:27	

---

Data Download Date/Time: 11/25/2019 05:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BASFORD,NICOLE,M  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: BASFORD,NICOLE  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 10/17/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/21/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 11/19/2019 20:39:18

---

Diagnostics Check	OK	20:32
Air Blank	0.000	20:33
Control Test	0.081	20:33
Air Blank	0.000	20:34
Subject Sample # 1	0.224	20:34
Breath Volume	1.894	20:34
Air Blank	0.000	20:35
Air Blank	0.000	20:37
Subject Sample # 2	0.220	20:37
Breath Volume	2.152	20:37
Air Blank	0.000	20:38
Control Test	0.079	20:38
Air Blank	0.000	20:39
Diagnostics Check	OK	20:39

---

Data Download Date/Time: 11/27/2019 12:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BASFORD,NICOLE,M  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: BASFORD,NICOLE  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 10/17/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/21/2021  
VIOLATION CODE: DUI  
ARREST TIME: 15:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:20

SUBJECT'S LAST NAME: ARTHUR

DATE/TIME OF TEST: 11/22/2019 15:57:31

---

Diagnostics Check	OK	15:51
Air Blank	0.000	15:51
Control Test	0.080	15:52
Air Blank	0.000	15:52
Subject Sample # 1	0.127	15:53
Breath Volume	1.902	15:53
Air Blank	0.000	15:53
Air Blank	0.000	15:55
Subject Sample # 2	0.132	15:56
Breath Volume	1.753	15:56
Air Blank	0.000	15:56
Control Test	0.079	15:56
Air Blank	0.000	15:57
Diagnostics Check	OK	15:57

---

Data Download Date/Time: 11/27/2019 12:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000901

INSTRUMENT REGISTERED TO: FFWCC TAMPA  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PULASKI,DAMON,J  
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)  
ARRESTING OFFICER: PULASKI,DAMON  
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)  
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A1  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 15:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:10

SUBJECT'S LAST NAME: SINGLETON

DATE/TIME OF TEST: 11/17/2019 16:12:10

---

Diagnostics Check	OK	16:07	
Air Blank	0.000	16:08	
Control Test	0.081	16:08	
Air Blank	0.000	16:09	
Subject Sample # 1	0.126	16:10	
Breath Volume	2.621	16:10	
Air Blank		16:11	Purge Fail
Air Blank	0.000	16:12	

---

Data Download Date/Time: 12/13/2019 13:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000901

INSTRUMENT REGISTERED TO: FFWCC TAMPA  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PULASKI,DAMON,J  
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)  
ARRESTING OFFICER: PULASKI,DAMON  
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)  
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A1  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 15:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:10

SUBJECT'S LAST NAME: SINGLETON

DATE/TIME OF TEST: 11/17/2019 16:22:59

---

Diagnostics Check	OK	16:15	
Air Blank	0.000	16:15	
Control Test	0.081	16:15	
Air Blank	0.000	16:16	
Subject Sample	# 1 0.110	16:19	Volume Not Met
Breath Volume	0.804	16:19	
Air Blank	0.000	16:20	
Air Blank	0.000	16:22	
Subject Sample	# 2	16:22	RFI Detect
Breath Volume	0.000	16:22	
Air Blank		16:22	RFI Detect
Air Blank	0.000	16:22	

---

Data Download Date/Time: 12/13/2019 13:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000901

INSTRUMENT REGISTERED TO: FFWCC TAMPA  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PULASKI,DAMON,J  
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)  
ARRESTING OFFICER: PULASKI,DAMON  
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)  
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A1  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 15:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:10

SUBJECT'S LAST NAME: SINGLETON

DATE/TIME OF TEST: 11/17/2019 16:33:41

---

Diagnostics Check	OK	16:26
Air Blank	0.000	16:27
Control Test	0.081	16:27
Air Blank	0.000	16:28
Subject Sample # 1	0.116	16:28
Breath Volume	1.484	16:28
Air Blank	0.000	16:29
Air Blank	0.000	16:31
Subject Sample # 2	0.109	16:32
Breath Volume	2.078	16:32
Air Blank	0.000	16:32
Control Test	0.080	16:33
Air Blank	0.000	16:33
Diagnostics Check	OK	16:33

---

Data Download Date/Time: 12/13/2019 13:34



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000901

INSTRUMENT REGISTERED TO: FFWCC TAMPA  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PULASKI,DAMON,J  
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)  
ARRESTING OFFICER: CONNELL,JACOB  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A1  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: DOERING

DATE/TIME OF TEST: 11/24/2019 04:33:10

---

Diagnostics Check	OK	04:30	
Air Blank	0.000	04:30	
Control Test	0.081	04:31	
Air Blank	0.000	04:31	
Subject Sample	# 1	04:31	Subject Test Refused
Breath Volume	0.000	04:31	
Air Blank	0.000	04:32	
Control Test	0.080	04:32	
Air Blank	0.000	04:33	
Diagnostics Check	OK	04:33	

---

Data Download Date/Time: 12/13/2019 13:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000902

INSTRUMENT REGISTERED TO: FFWCC TAMPA  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILKINS,MALACHI,C  
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)  
ARRESTING OFFICER: WILKINS,MALACHI  
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)  
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: BUI  
ARREST TIME: 19:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:23

SUBJECT'S LAST NAME: LENNON

DATE/TIME OF TEST: 11/30/2019 20:01:26

---

Diagnostics Check	OK	19:59	
Air Blank	0.000	19:59	
Control Test	0.079	20:00	
Air Blank		20:00	Purge Fail
Air Blank	0.000	20:01	

---

Data Download Date/Time: 12/13/2019 13:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000902

INSTRUMENT REGISTERED TO: FFWCC TAMPA  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILKINS,MALACHI,C  
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)  
ARRESTING OFFICER: WILKINS,MALACHI  
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)  
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: BUI  
ARREST TIME: 19:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:23

SUBJECT'S LAST NAME: LENNON

DATE/TIME OF TEST: 11/30/2019 20:07:52

---

Diagnostics Check	OK	20:04	
Air Blank	0.000	20:05	
Control Test	0.079	20:05	
Air Blank	0.000	20:06	
Subject Sample	# 1	20:06	Subject Test Refused
Breath Volume	0.000	20:06	
Air Blank	0.000	20:06	
Control Test	0.080	20:07	
Air Blank	0.000	20:07	
Diagnostics Check	OK	20:07	

---

Data Download Date/Time: 12/13/2019 13:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROWE,BRANDON,D  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: ROWE,BRANDON  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 10/21/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 867139  
DRY GAS EXPIRY DATE: 11/15/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: ROWLAND

DATE/TIME OF TEST: 11/01/2019 00:56:13

---

Diagnostics Check	OK	00:50
Air Blank	0.000	00:50
Control Test	0.078	00:50
Air Blank	0.000	00:51
Subject Sample # 1	0.000	00:51
Breath Volume	2.074	00:51
Air Blank	0.000	00:52
Air Blank	0.000	00:54
Subject Sample # 2	0.000	00:54
Breath Volume	1.636	00:54
Air Blank	0.000	00:55
Control Test	0.078	00:55
Air Blank	0.000	00:56
Diagnostics Check	OK	00:56

---

Data Download Date/Time: 11/18/2019 08:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROWE,BRANDON,D  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: ROWE,BRANDON  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 10/21/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 867139  
DRY GAS EXPIRY DATE: 11/15/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: MILLS

DATE/TIME OF TEST: 11/09/2019 23:02:03

---

Diagnostics Check	OK	22:51	
Air Blank	0.000	22:52	
Control Test	0.079	22:52	
Air Blank	0.000	22:52	
Subject Sample # 1	0.000	22:55	
Breath Volume	1.421	22:55	
Air Blank	0.000	22:55	
Air Blank	0.000	22:57	
Subject Sample # 2		23:00	Subject Test Refused
Breath Volume	0.304	23:00	
Air Blank	0.000	23:01	
Control Test	0.079	23:01	
Air Blank	0.000	23:01	
Diagnostics Check	OK	23:02	

---

Data Download Date/Time: 11/18/2019 08:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROWE,BRANDON,D  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: ROWE,BRANDON  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 10/21/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 867139  
DRY GAS EXPIRY DATE: 11/15/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: ANGIOLILLO

DATE/TIME OF TEST: 11/10/2019 03:08:00

---

Diagnostics Check	OK	03:01
Air Blank	0.000	03:02
Control Test	0.080	03:02
Air Blank	0.000	03:03
Subject Sample # 1	0.175	03:03
Breath Volume	1.886	03:03
Air Blank	0.000	03:04
Air Blank	0.000	03:06
Subject Sample # 2	0.172	03:06
Breath Volume	1.511	03:06
Air Blank	0.000	03:07
Control Test	0.079	03:07
Air Blank	0.000	03:07
Diagnostics Check	OK	03:08

---

Data Download Date/Time: 11/18/2019 08:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROWE,BRANDON,D  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: ROWE,BRANDON  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 10/21/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 867139  
DRY GAS EXPIRY DATE: 11/15/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 11/11/2019 00:50:42

---

Diagnostics Check	OK	00:41	
Air Blank	0.000	00:42	
Control Test	0.080	00:42	
Air Blank	0.000	00:43	
Subject Sample	# 1	00:46	No Sample Provided
Breath Volume	0.000	00:46	
Air Blank	0.000	00:46	
Air Blank	0.000	00:48	
Subject Sample	# 2	00:49	Subject Test Refused
Breath Volume	0.000	00:49	
Air Blank	0.000	00:49	
Control Test	0.079	00:50	
Air Blank	0.000	00:50	
Diagnostics Check	OK	00:50	

---

Data Download Date/Time: 11/18/2019 08:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROWE,BRANDON,D  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: ROWE,BRANDON  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 11/18/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 867139  
DRY GAS EXPIRY DATE: 11/15/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: AGUINAGA SEGOVIA

DATE/TIME OF TEST: 11/19/2019 22:32:08

---

Diagnostics Check	OK	22:25
Air Blank	0.000	22:26
Control Test	0.079	22:26
Air Blank	0.000	22:27
Subject Sample # 1	0.080	22:27
Breath Volume	5.046	22:27
Air Blank	0.000	22:28
Air Blank	0.000	22:30
Subject Sample # 2	0.078	22:30
Breath Volume	5.363	22:30
Air Blank	0.000	22:31
Control Test	0.079	22:31
Air Blank	0.000	22:31
Diagnostics Check	OK	22:32

---

Data Download Date/Time: 12/17/2019 09:22



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROWE,BRANDON,D  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: ROWE,BRANDON  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 11/18/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 867139  
DRY GAS EXPIRY DATE: 11/15/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: CORTEZ GUZMAN

DATE/TIME OF TEST: 11/23/2019 01:04:46

---

Diagnostics Check	OK	00:58
Air Blank	0.000	00:59
Control Test	0.080	00:59
Air Blank	0.000	00:59
Subject Sample # 1	0.191	01:00
Breath Volume	1.699	01:00
Air Blank	0.000	01:01
Air Blank	0.000	01:03
Subject Sample # 2	0.186	01:03
Breath Volume	1.515	01:03
Air Blank	0.000	01:03
Control Test	0.079	01:04
Air Blank	0.000	01:04
Diagnostics Check	OK	01:04

---

Data Download Date/Time: 12/17/2019 09:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROWE,BRANDON,D  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: ROWE,BRANDON  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 11/18/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 867139  
DRY GAS EXPIRY DATE: 11/15/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: FULCHER

DATE/TIME OF TEST: 11/24/2019 23:41:08

---

Diagnostics Check	OK	23:34
Air Blank	0.000	23:35
Control Test	0.080	23:35
Air Blank	0.000	23:36
Subject Sample # 1	0.195	23:36
Breath Volume	1.750	23:36
Air Blank	0.000	23:37
Air Blank	0.000	23:39
Subject Sample # 2	0.186	23:39
Breath Volume	2.035	23:39
Air Blank	0.000	23:40
Control Test	0.080	23:40
Air Blank	0.000	23:41
Diagnostics Check	OK	23:41

---

Data Download Date/Time: 12/17/2019 09:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000908

INSTRUMENT REGISTERED TO: PORT ST JOE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VAN HEERDEN, TYRON, C  
BREATH TEST OPERATOR AGENCY: PORT ST. JOE PD  
ARRESTING OFFICER: VAN HEERDEN, TYRON  
ARRESTING OFFICER AGENCY: PORT ST. JOE PD  
LAST AGENCY INSPECTION: 10/01/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A1  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 15:24  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 16:28

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 11/24/2019 17:13:22

---

Diagnostics Check		OK	16:59	
Air Blank		0.000	16:59	
Control Test		0.078	16:59	
Air Blank		0.000	17:00	
Subject Sample	# 1	0.136	17:03	Volume Not Met
Breath Volume		0.210	17:03	
Air Blank		0.000	17:04	
Air Blank		0.000	17:06	
Subject Sample	# 2	0.134	17:08	
Breath Volume		1.710	17:08	
Air Blank		0.000	17:09	
Air Blank		0.000	17:11	
Subject Sample	# 3	0.124	17:11	
Breath Volume		1.816	17:11	
Air Blank		0.000	17:12	
Control Test		0.078	17:12	
Air Blank		0.000	17:13	
Diagnostics Check		OK	17:13	

---

Data Download Date/Time: 11/25/2019 20:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000908

INSTRUMENT REGISTERED TO: PORT ST JOE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DICKEY, WILLIAM, Q  
BREATH TEST OPERATOR AGENCY: PORT ST. JOE PD  
ARRESTING OFFICER: YOUNG, PAUL  
ARRESTING OFFICER AGENCY: GULF COUNTY SO  
LAST AGENCY INSPECTION: 11/25/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A1  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: QUAY

DATE/TIME OF TEST: 11/30/2019 00:17:08

---

Diagnostics Check	OK	00:02	
Air Blank	0.000	00:03	
Control Test	0.078	00:03	
Air Blank	0.000	00:03	
Subject Sample # 1	0.134	00:07	
Breath Volume	1.394	00:07	
Air Blank	0.000	00:07	
Air Blank	0.000	00:09	
Subject Sample # 2	0.142	00:12	Volume Not Met
Breath Volume	1.031	00:12	
Air Blank	0.000	00:13	
Air Blank	0.000	00:15	
Subject Sample # 3	0.153	00:15	
Breath Volume	2.152	00:15	
Air Blank	0.000	00:16	
Control Test	0.078	00:16	
Air Blank	0.000	00:17	
Diagnostics Check	OK	00:17	

---

Data Download Date/Time: 12/15/2019 19:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMYRNIOS,GEORGE,T  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: SMYRNIOS,GEORGE  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 15:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:24

SUBJECT'S LAST NAME: POSTLEWAIT

DATE/TIME OF TEST: 11/01/2019 17:51:01

---

Diagnostics Check	OK	17:44
Air Blank	0.000	17:45
Control Test	0.079	17:45
Air Blank	0.000	17:46
Subject Sample # 1	0.240	17:46
Breath Volume	2.375	17:46
Air Blank	0.000	17:47
Air Blank	0.000	17:49
Subject Sample # 2	0.242	17:49
Breath Volume	1.996	17:49
Air Blank	0.000	17:50
Control Test	0.079	17:50
Air Blank	0.000	17:50
Diagnostics Check	OK	17:51

---

Data Download Date/Time: 11/20/2019 10:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH, JOSIAH, R  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: SMITH, JOSIAH  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 17:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:31

SUBJECT'S LAST NAME: BEHEN

DATE/TIME OF TEST: 11/03/2019 19:00:44

---

Diagnostics Check	OK	18:54
Air Blank	0.000	18:54
Control Test	0.079	18:54
Air Blank	0.000	18:55
Subject Sample # 1	0.127	18:56
Breath Volume	1.734	18:56
Air Blank	0.000	18:56
Air Blank	0.000	18:58
Subject Sample # 2	0.124	18:59
Breath Volume	1.617	18:59
Air Blank	0.000	18:59
Control Test	0.079	19:00
Air Blank	0.000	19:00
Diagnostics Check	OK	19:00

---

Data Download Date/Time: 11/20/2019 10:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BEINER,RYAN,T  
BREATH TEST OPERATOR AGENCY: Fort Myers PD  
ARRESTING OFFICER: BEINER,RYAN  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:32

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 11/03/2019 20:02:08

---

Diagnostics Check	OK	19:55
Air Blank	0.000	19:56
Control Test	0.079	19:56
Air Blank	0.000	19:57
Subject Sample # 1	0.218	19:57
Breath Volume	1.472	19:57
Air Blank	0.000	19:58
Air Blank	0.000	20:00
Subject Sample # 2	0.204	20:00
Breath Volume	1.476	20:00
Air Blank	0.000	20:01
Control Test	0.079	20:01
Air Blank	0.000	20:02
Diagnostics Check	OK	20:02

---

Data Download Date/Time: 11/20/2019 10:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH, JOSIAH, R  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: SELIMOVIC, A  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:47

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 11/03/2019 22:19:30

---

Diagnostics Check	OK	22:12
Air Blank	0.000	22:13
Control Test	0.079	22:13
Air Blank	0.000	22:13
Subject Sample # 1	0.183	22:15
Breath Volume	1.730	22:15
Air Blank	0.000	22:15
Air Blank	0.000	22:17
Subject Sample # 2	0.190	22:18
Breath Volume	2.046	22:18
Air Blank	0.000	22:18
Control Test	0.079	22:18
Air Blank	0.000	22:19
Diagnostics Check	OK	22:19

---

Data Download Date/Time: 11/20/2019 10:42



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WAY,JOSEPH,T  
BREATH TEST OPERATOR AGENCY: Fort Myers PD  
ARRESTING OFFICER: WAY,JOSEPH  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 13:58  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 14:21

SUBJECT'S LAST NAME: PATTON

DATE/TIME OF TEST: 11/05/2019 15:03:47

---

Diagnostics Check	OK	14:57
Air Blank	0.000	14:58
Control Test	0.079	14:58
Air Blank	0.000	14:58
Subject Sample # 1	0.000	14:59
Breath Volume	2.183	14:59
Air Blank	0.000	14:59
Air Blank	0.000	15:01
Subject Sample # 2	0.000	15:02
Breath Volume	1.910	15:02
Air Blank	0.000	15:02
Control Test	0.079	15:03
Air Blank	0.000	15:03
Diagnostics Check	OK	15:03

---

Data Download Date/Time: 11/20/2019 10:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WINCELOWICZ,LARRY,R  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: WINCELOWICZ,LARRY  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: LANYON

DATE/TIME OF TEST: 11/07/2019 23:09:42

---

Diagnostics Check	OK	23:03
Air Blank	0.000	23:03
Control Test	0.080	23:04
Air Blank	0.000	23:04
Subject Sample # 1	0.115	23:05
Breath Volume	1.984	23:05
Air Blank	0.000	23:05
Air Blank	0.000	23:07
Subject Sample # 2	0.110	23:08
Breath Volume	1.488	23:08
Air Blank	0.000	23:08
Control Test	0.079	23:09
Air Blank	0.000	23:09
Diagnostics Check	OK	23:09

---

Data Download Date/Time: 11/20/2019 10:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WAY,JOSEPH,T  
BREATH TEST OPERATOR AGENCY: Fort Myers PD  
ARRESTING OFFICER: WAY,JOSEPH  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 16:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:32

SUBJECT'S LAST NAME: GAVERE

DATE/TIME OF TEST: 11/08/2019 17:01:45

---

Diagnostics Check	OK	16:55
Air Blank	0.000	16:55
Control Test	0.080	16:56
Air Blank	0.000	16:56
Subject Sample # 1	0.099	16:57
Breath Volume	1.156	16:57
Air Blank	0.000	16:57
Air Blank	0.000	16:59
Subject Sample # 2	0.100	17:00
Breath Volume	2.089	17:00
Air Blank	0.000	17:00
Control Test	0.080	17:01
Air Blank	0.000	17:01
Diagnostics Check	OK	17:01

---

Data Download Date/Time: 11/20/2019 10:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KIRK,PATRICK,T  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: KIRK,PATRICK  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: CALDWELL

DATE/TIME OF TEST: 11/10/2019 04:04:55

---

Diagnostics Check	OK	03:58
Air Blank	0.000	03:59
Control Test	0.079	03:59
Air Blank	0.000	04:00
Subject Sample # 1	0.141	04:00
Breath Volume	2.074	04:00
Air Blank	0.000	04:01
Air Blank	0.000	04:02
Subject Sample # 2	0.142	04:03
Breath Volume	2.343	04:03
Air Blank	0.000	04:04
Control Test	0.079	04:04
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

---

Data Download Date/Time: 11/20/2019 10:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TISDALE,KATIE,E  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: DAVOLT,TYLER  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 15:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:35

SUBJECT'S LAST NAME: SANDERS

DATE/TIME OF TEST: 11/11/2019 17:12:58

---

Diagnostics Check	OK	17:06
Air Blank	0.000	17:07
Control Test	0.079	17:07
Air Blank	0.000	17:08
Subject Sample # 1	0.000	17:08
Breath Volume	2.613	17:08
Air Blank	0.000	17:09
Air Blank	0.000	17:11
Subject Sample # 2	0.000	17:11
Breath Volume	1.996	17:11
Air Blank	0.000	17:12
Control Test	0.079	17:12
Air Blank	0.000	17:12
Diagnostics Check	OK	17:12

---

Data Download Date/Time: 11/20/2019 10:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TISDALE,KATIE,E  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: CAMEJO,JASMIN  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:38  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: SHIRES

DATE/TIME OF TEST: 11/12/2019 02:34:53

---

Diagnostics Check	OK	02:28
Air Blank	0.000	02:28
Control Test	0.080	02:28
Air Blank	0.000	02:29
Subject Sample # 1	0.268	02:30
Breath Volume	1.914	02:30
Air Blank	0.000	02:30
Air Blank	0.000	02:32
Subject Sample # 2	0.263	02:33
Breath Volume	2.089	02:33
Air Blank	0.000	02:33
Control Test	0.078	02:34
Air Blank	0.000	02:34
Diagnostics Check	OK	02:34

---

Data Download Date/Time: 11/20/2019 10:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BEINER,RYAN,T  
BREATH TEST OPERATOR AGENCY: Fort Myers PD  
ARRESTING OFFICER: BEINER,RYAN  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: SANTOS SALINAS

DATE/TIME OF TEST: 11/12/2019 03:50:22

---

Diagnostics Check		OK	03:44
Air Blank		0.000	03:44
Control Test		0.079	03:44
Air Blank		0.000	03:45
Subject Sample	# 1	0.208	03:45
Breath Volume		1.777	03:45
Air Blank		0.000	03:46
Air Blank		0.000	03:48
Subject Sample	# 2	0.188	03:48
Breath Volume		1.109	03:48
Air Blank		0.000	03:49
Control Test		0.079	03:49
Air Blank		0.000	03:50
Diagnostics Check		OK	03:50

---

Data Download Date/Time: 11/20/2019 10:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH,JOSIAH,R  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: SMITH,JOSIAH  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: QUINTERO ARANGO

DATE/TIME OF TEST: 11/13/2019 03:05:45

---

Diagnostics Check	OK	02:58
Air Blank	0.000	02:59
Control Test	0.079	02:59
Air Blank	0.000	02:59
Subject Sample # 1	0.195	03:01
Breath Volume	2.828	03:01
Air Blank	0.000	03:02
Air Blank	0.000	03:03
Subject Sample # 2	0.192	03:04
Breath Volume	1.992	03:04
Air Blank	0.000	03:04
Control Test	0.078	03:05
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

---

Data Download Date/Time: 11/20/2019 10:42



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HALL,CHRISTINE,  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: BAIR,JAMES  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:36  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: FOX

DATE/TIME OF TEST: 11/15/2019 23:46:48

---

Diagnostics Check		OK	23:33	
Air Blank		0.000	23:33	
Control Test		0.079	23:34	
Air Blank		0.000	23:34	
Subject Sample	# 1	0.196	23:37	Volume Not Met
Breath Volume		0.148	23:37	
Air Blank		0.000	23:38	
Air Blank		0.000	23:40	
Subject Sample	# 2	0.174	23:41	No .020 agreement
Breath Volume		1.265	23:41	
Air Blank		0.000	23:41	
Air Blank		0.000	23:43	
Subject Sample	# 3	0.203	23:45	No .020 agreement
Breath Volume		1.273	23:45	
Air Blank		0.000	23:45	
Control Test		0.079	23:46	
Air Blank		0.000	23:46	
Diagnostics Check		OK	23:46	

---

Data Download Date/Time: 11/20/2019 10:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HALL,CHRISTINE,  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: BAIR,JAMES  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:36  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: FOX

DATE/TIME OF TEST: 11/16/2019 00:01:02

---

Diagnostics Check	OK	23:54
Air Blank	0.000	23:54
Control Test	0.080	23:55
Air Blank	0.000	23:55
Subject Sample # 1	0.187	23:56
Breath Volume	1.875	23:56
Air Blank	0.000	23:56
Air Blank	0.000	23:58
Subject Sample # 2	0.194	23:59
Breath Volume	2.015	23:59
Air Blank	0.000	00:00
Control Test	0.079	00:00
Air Blank	0.000	00:00
Diagnostics Check	OK	00:01

---

Data Download Date/Time: 11/20/2019 10:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTINS,KYLE,L  
BREATH TEST OPERATOR AGENCY: Fort Myers PD  
ARRESTING OFFICER: MARTINS,KYLE  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: LLAMAS MARTINEZ

DATE/TIME OF TEST: 11/16/2019 05:14:03

---

Diagnostics Check		OK	05:07
Air Blank		0.000	05:08
Control Test		0.080	05:08
Air Blank		0.000	05:09
Subject Sample	# 1	0.203	05:09
Breath Volume		2.652	05:09
Air Blank		0.000	05:10
Air Blank		0.000	05:12
Subject Sample	# 2	0.201	05:12
Breath Volume		2.964	05:12
Air Blank		0.000	05:13
Control Test		0.079	05:13
Air Blank		0.000	05:13
Diagnostics Check		OK	05:14

---

Data Download Date/Time: 11/20/2019 10:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTINS,KYLE,L  
BREATH TEST OPERATOR AGENCY: Fort Myers PD  
ARRESTING OFFICER: SUTTON,BRYAN  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 05:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:35

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 11/16/2019 07:03:28

---

Diagnostics Check	OK	06:57
Air Blank	0.000	06:57
Control Test	0.080	06:57
Air Blank	0.000	06:58
Subject Sample # 1	0.144	06:58
Breath Volume	5.187	06:58
Air Blank	0.000	06:59
Air Blank	0.000	07:01
Subject Sample # 2	0.141	07:01
Breath Volume	4.390	07:01
Air Blank	0.000	07:02
Control Test	0.079	07:02
Air Blank	0.000	07:03
Diagnostics Check	OK	07:03

---

Data Download Date/Time: 11/20/2019 10:42



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TISDALE,KATIE,E  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: CAIVANO,STACEY  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: POELKER

DATE/TIME OF TEST: 11/17/2019 03:12:37

---

Diagnostics Check	OK	03:06
Air Blank	0.000	03:06
Control Test	0.080	03:07
Air Blank	0.000	03:07
Subject Sample # 1	0.170	03:08
Breath Volume	2.457	03:08
Air Blank	0.000	03:08
Air Blank	0.000	03:10
Subject Sample # 2	0.169	03:11
Breath Volume	2.214	03:11
Air Blank	0.000	03:11
Control Test	0.079	03:12
Air Blank	0.000	03:12
Diagnostics Check	OK	03:12

---

Data Download Date/Time: 11/20/2019 10:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TISDALE,KATIE,E  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: KUPKA,RIKKI  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 18:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:54

SUBJECT'S LAST NAME: O'TOOLE

DATE/TIME OF TEST: 11/17/2019 19:28:43

---

Diagnostics Check	OK	19:22
Air Blank	0.000	19:22
Control Test	0.080	19:22
Air Blank	0.000	19:23
Subject Sample # 1	0.267	19:24
Breath Volume	1.812	19:24
Air Blank	0.000	19:24
Air Blank	0.000	19:26
Subject Sample # 2	0.266	19:27
Breath Volume	2.109	19:27
Air Blank	0.000	19:27
Control Test	0.079	19:28
Air Blank	0.000	19:28
Diagnostics Check	OK	19:28

---

Data Download Date/Time: 11/20/2019 10:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TISDALE,KATIE,E  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: GONZALES,DAVID  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: GASPAR

DATE/TIME OF TEST: 11/17/2019 23:45:21

---

Diagnostics Check	OK	23:39
Air Blank	0.000	23:39
Control Test	0.081	23:39
Air Blank	0.000	23:40
Subject Sample # 1	0.187	23:40
Breath Volume	2.304	23:40
Air Blank	0.000	23:41
Air Blank	0.000	23:43
Subject Sample # 2	0.185	23:43
Breath Volume	2.292	23:43
Air Blank	0.000	23:44
Control Test	0.080	23:44
Air Blank	0.000	23:45
Diagnostics Check	OK	23:45

---

Data Download Date/Time: 11/20/2019 10:42



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WAY,JOSEPH,T  
BREATH TEST OPERATOR AGENCY: Fort Myers PD  
ARRESTING OFFICER: WAY,JOSEPH  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 15:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: BESTONSO

DATE/TIME OF TEST: 11/18/2019 16:37:10

---

Diagnostics Check	OK	16:30
Air Blank	0.000	16:31
Control Test	0.080	16:31
Air Blank	0.000	16:32
Subject Sample # 1	0.000	16:32
Breath Volume	2.972	16:32
Air Blank	0.000	16:33
Air Blank	0.000	16:35
Subject Sample # 2	0.000	16:35
Breath Volume	3.527	16:35
Air Blank	0.000	16:36
Control Test	0.080	16:36
Air Blank	0.000	16:37
Diagnostics Check	OK	16:37

---

Data Download Date/Time: 11/20/2019 10:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYERS,JUSTIN,L  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: MYERS,JUSTIN  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:37  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: FUSS

DATE/TIME OF TEST: 11/20/2019 02:15:06

---

Diagnostics Check	OK	02:07
Air Blank	0.000	02:08
Control Test	0.080	02:08
Air Blank	0.000	02:08
Subject Sample # 1	0.140	02:10
Breath Volume	1.976	02:10
Air Blank	0.000	02:10
Air Blank	0.000	02:12
Subject Sample # 2	0.138	02:13
Breath Volume	2.496	02:13
Air Blank	0.000	02:14
Control Test	0.079	02:14
Air Blank	0.000	02:15
Diagnostics Check	OK	02:15

---

Data Download Date/Time: 11/20/2019 10:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH,JOSIAH,R  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: SMITH,JOSIAH  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 11/20/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:08  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: POLITE

DATE/TIME OF TEST: 11/22/2019 01:57:47

---

Diagnostics Check	OK	01:43	
Air Blank	0.000	01:43	
Control Test	0.080	01:44	
Air Blank	0.000	01:44	
Subject Sample	# 1 0.172	01:46	
Breath Volume	1.312	01:46	
Air Blank	0.000	01:46	
Air Blank	0.000	01:48	
Subject Sample	# 2 0.172	01:51	Volume Not Met
Breath Volume	0.738	01:51	
Air Blank	0.000	01:52	
Air Blank	0.000	01:54	
Subject Sample	# 3 0.185	01:56	
Breath Volume	1.425	01:56	
Air Blank	0.000	01:56	
Control Test	0.078	01:57	
Air Blank	0.000	01:57	
Diagnostics Check	OK	01:57	

---

Data Download Date/Time: 12/11/2019 12:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARDIN, JEREMY, L  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: BECKOS, JAMES  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 11/20/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 05:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:02

SUBJECT'S LAST NAME: FREDRICKSON

DATE/TIME OF TEST: 11/23/2019 06:28:41

---

Diagnostics Check	OK	06:23	
Air Blank	0.000	06:23	
Control Test	0.080	06:23	
Air Blank	0.000	06:24	
Subject Sample	# 1 0.182	06:24	
Breath Volume	3.007	06:24	
Air Blank	0.000	06:25	
Air Blank	0.000	06:27	
Subject Sample	# 2	06:27	Subject Test Refused
Breath Volume	0.000	06:27	
Air Blank	0.000	06:27	
Control Test	0.078	06:28	
Air Blank	0.000	06:28	
Diagnostics Check	OK	06:28	

---

Data Download Date/Time: 12/11/2019 12:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHAEFER, KENDALL, A  
BREATH TEST OPERATOR AGENCY: Fort Myers PD  
ARRESTING OFFICER: BRITO, DAVID  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 11/20/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 11:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:40

SUBJECT'S LAST NAME: MAYOLI PEREZ

DATE/TIME OF TEST: 11/23/2019 14:09:07

---

Diagnostics Check	OK	14:03
Air Blank	0.000	14:03
Control Test	0.079	14:04
Air Blank	0.000	14:04
Subject Sample # 1	0.106	14:04
Breath Volume	2.097	14:04
Air Blank	0.000	14:05
Air Blank	0.000	14:07
Subject Sample # 2	0.103	14:07
Breath Volume	2.257	14:07
Air Blank	0.000	14:08
Control Test	0.079	14:08
Air Blank	0.000	14:09
Diagnostics Check	OK	14:09

---

Data Download Date/Time: 12/11/2019 12:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CANFIELD 18096,A.,B  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: COLWELL 18-24,C.  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 11/20/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:08  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:23

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 11/23/2019 22:22:55

---

Diagnostics Check	OK	22:15
Air Blank	0.000	22:16
Control Test	0.080	22:16
Air Blank	0.000	22:17
Subject Sample # 1	0.197	22:18
Breath Volume	1.363	22:18
Air Blank	0.000	22:18
Air Blank	0.000	22:20
Subject Sample # 2	0.189	22:21
Breath Volume	1.648	22:21
Air Blank	0.000	22:22
Control Test	0.079	22:22
Air Blank	0.000	22:22
Diagnostics Check	OK	22:22

---

Data Download Date/Time: 12/11/2019 12:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYERS,JUSTIN,L  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: MYERS,JUSTIN  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 11/20/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:41  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: WILKENS

DATE/TIME OF TEST: 11/24/2019 03:02:36

---

Diagnostics Check	OK	02:56
Air Blank	0.000	02:56
Control Test	0.079	02:57
Air Blank	0.000	02:57
Subject Sample # 1	0.158	02:58
Breath Volume	1.957	02:58
Air Blank	0.000	02:58
Air Blank	0.000	03:00
Subject Sample # 2	0.158	03:01
Breath Volume	2.273	03:01
Air Blank	0.000	03:01
Control Test	0.079	03:02
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

---

Data Download Date/Time: 12/11/2019 12:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LUX UM,JUAN,C  
BREATH TEST OPERATOR AGENCY: Fort Myers PD  
ARRESTING OFFICER: TOWNSEND,ANTHONY  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 11/20/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 16:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:09

SUBJECT'S LAST NAME: LUX UM

DATE/TIME OF TEST: 11/24/2019 17:42:51

---

Diagnostics Check	OK	17:36
Air Blank	0.000	17:37
Control Test	0.079	17:37
Air Blank	0.000	17:37
Subject Sample # 1	0.206	17:38
Breath Volume	1.695	17:38
Air Blank	0.000	17:39
Air Blank	0.000	17:40
Subject Sample # 2	0.191	17:41
Breath Volume	1.855	17:41
Air Blank	0.000	17:41
Control Test	0.078	17:42
Air Blank	0.000	17:42
Diagnostics Check	OK	17:42

---

Data Download Date/Time: 12/11/2019 12:39



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROMAN,A,C  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: ELSCHLAGER,A  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 11/20/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: CHEW

DATE/TIME OF TEST: 11/25/2019 05:59:26

---

Diagnostics Check	OK	05:52
Air Blank	0.000	05:53
Control Test	0.081	05:53
Air Blank	0.000	05:54
Subject Sample # 1	0.000	05:54
Breath Volume	2.007	05:54
Air Blank	0.000	05:54
Air Blank	0.000	05:56
Subject Sample # 2	0.000	05:58
Breath Volume	1.367	05:58
Air Blank	0.000	05:58
Control Test	0.080	05:58
Air Blank	0.000	05:59
Diagnostics Check	OK	05:59

---

Data Download Date/Time: 12/11/2019 12:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYERS,JUSTIN,L  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: MYERS,JUSTIN  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 11/20/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:23

SUBJECT'S LAST NAME: SCHROEDER

DATE/TIME OF TEST: 11/27/2019 22:46:23

---

Diagnostics Check	OK	22:43	
Air Blank	0.000	22:43	
Control Test	0.080	22:44	
Air Blank	0.000	22:44	
Subject Sample	# 1	22:45	Subject Test Refused
Breath Volume	0.000	22:45	
Air Blank	0.000	22:45	
Control Test	0.080	22:45	
Air Blank	0.000	22:46	
Diagnostics Check	OK	22:46	

---

Data Download Date/Time: 12/11/2019 12:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYERS,JUSTIN,L  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: THOMAS,DENNIS  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 11/20/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:23

SUBJECT'S LAST NAME: GRAY

DATE/TIME OF TEST: 11/27/2019 22:55:45

---

Diagnostics Check	OK	22:49
Air Blank	0.000	22:49
Control Test	0.079	22:49
Air Blank	0.000	22:50
Subject Sample # 1	0.127	22:51
Breath Volume	1.535	22:51
Air Blank	0.000	22:51
Air Blank	0.000	22:53
Subject Sample # 2	0.126	22:54
Breath Volume	2.535	22:54
Air Blank	0.000	22:54
Control Test	0.079	22:55
Air Blank	0.000	22:55
Diagnostics Check	OK	22:55

---

Data Download Date/Time: 12/11/2019 12:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYERS,JUSTIN,L  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: MYERS,JUSTIN  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 11/20/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:23

SUBJECT'S LAST NAME: SCHROEDER

DATE/TIME OF TEST: 11/27/2019 23:48:00

---

Diagnostics Check	OK	23:41
Air Blank	0.000	23:41
Control Test	0.079	23:41
Air Blank	0.000	23:42
Subject Sample # 1	0.215	23:43
Breath Volume	2.765	23:43
Air Blank	0.000	23:43
Air Blank	0.000	23:45
Subject Sample # 2	0.207	23:46
Breath Volume	2.730	23:46
Air Blank	0.000	23:47
Control Test	0.080	23:47
Air Blank	0.000	23:47
Diagnostics Check	OK	23:48

---

Data Download Date/Time: 12/11/2019 12:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RUSH,DANIEL,C  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: RUSH,DANIEL  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 11/20/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: NOGUEROL-PEREZ

DATE/TIME OF TEST: 11/29/2019 01:06:23

---

Diagnostics Check	OK	00:59
Air Blank	0.000	01:00
Control Test	0.080	01:00
Air Blank	0.000	01:01
Subject Sample # 1	0.139	01:01
Breath Volume	2.332	01:01
Air Blank	0.000	01:02
Air Blank	0.000	01:04
Subject Sample # 2	0.142	01:04
Breath Volume	3.085	01:04
Air Blank	0.000	01:05
Control Test	0.079	01:05
Air Blank	0.000	01:06
Diagnostics Check	OK	01:06

---

Data Download Date/Time: 12/11/2019 12:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CANFIELD 18096,A.,B  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: SUCHKOVA 09089,N  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 11/20/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 11/29/2019 01:31:13

---

Diagnostics Check	OK	01:24
Air Blank	0.000	01:24
Control Test	0.080	01:24
Air Blank	0.000	01:25
Subject Sample # 1	0.207	01:26
Breath Volume	2.574	01:26
Air Blank	0.000	01:27
Air Blank	0.000	01:29
Subject Sample # 2	0.212	01:29
Breath Volume	1.566	01:29
Air Blank	0.000	01:30
Control Test	0.080	01:30
Air Blank	0.000	01:31
Diagnostics Check	OK	01:31

---

Data Download Date/Time: 12/11/2019 12:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH, JOSIAH, R  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: SMITH, JOSIAH  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 11/20/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:50  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: KOLAKOWSKI

DATE/TIME OF TEST: 11/30/2019 01:06:30

---

Diagnostics Check		OK	00:59
Air Blank		0.000	01:00
Control Test		0.080	01:00
Air Blank		0.000	01:01
Subject Sample	# 1	0.161	01:02
Breath Volume		1.957	01:02
Air Blank		0.000	01:02
Air Blank		0.000	01:04
Subject Sample	# 2	0.166	01:05
Breath Volume		2.125	01:05
Air Blank		0.000	01:05
Control Test		0.079	01:05
Air Blank		0.000	01:06
Diagnostics Check		OK	01:06

---

Data Download Date/Time: 12/11/2019 12:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROMAN,A,C  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: ROSALES,P  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 11/20/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 06:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:40

SUBJECT'S LAST NAME: MIRR

DATE/TIME OF TEST: 11/30/2019 07:06:28

---

Diagnostics Check	OK	07:00
Air Blank	0.000	07:01
Control Test	0.080	07:01
Air Blank	0.000	07:01
Subject Sample # 1	0.022	07:02
Breath Volume	2.054	07:02
Air Blank	0.000	07:02
Air Blank	0.000	07:04
Subject Sample # 2	0.022	07:05
Breath Volume	2.476	07:05
Air Blank	0.000	07:05
Control Test	0.079	07:05
Air Blank	0.000	07:06
Diagnostics Check	OK	07:06

---

Data Download Date/Time: 12/11/2019 12:40



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTINS,KYLE,L  
BREATH TEST OPERATOR AGENCY: Fort Myers PD  
ARRESTING OFFICER: AMADOR,RYAN  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 11/20/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:43

SUBJECT'S LAST NAME: GARCIA GARCIA

DATE/TIME OF TEST: 11/30/2019 22:32:48

---

Diagnostics Check	OK	22:26
Air Blank	0.000	22:27
Control Test	0.080	22:27
Air Blank	0.000	22:28
Subject Sample # 1	0.015	22:28
Breath Volume	2.957	22:28
Air Blank	0.000	22:29
Air Blank	0.000	22:30
Subject Sample # 2	0.014	22:31
Breath Volume	3.339	22:31
Air Blank	0.000	22:31
Control Test	0.080	22:32
Air Blank	0.000	22:32
Diagnostics Check	OK	22:32

---

Data Download Date/Time: 12/11/2019 12:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AVENDANO,JUAN,M  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: SERRAGO,JAMES  
ARRESTING OFFICER AGENCY: Sanibel PD  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:55  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:44

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 11/01/2019 22:12:22

---

Diagnostics Check	OK	22:05
Air Blank	0.000	22:06
Control Test	0.081	22:06
Air Blank	0.000	22:07
Subject Sample # 1	0.101	22:07
Breath Volume	1.527	22:07
Air Blank	0.000	22:08
Air Blank	0.000	22:10
Subject Sample # 2	0.102	22:10
Breath Volume	2.007	22:10
Air Blank	0.000	22:11
Control Test	0.081	22:11
Air Blank	0.000	22:12
Diagnostics Check	OK	22:12

---

Data Download Date/Time: 11/20/2019 10:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BELL,JOSEPH,G  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: GIZZI,VINCENT  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:52  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: JANS

DATE/TIME OF TEST: 11/04/2019 02:06:46

---

Diagnostics Check	OK	01:59
Air Blank	0.000	01:59
Control Test	0.081	02:00
Air Blank	0.000	02:00
Subject Sample # 1	0.124	02:01
Breath Volume	1.203	02:01
Air Blank	0.000	02:02
Air Blank	0.000	02:04
Subject Sample # 2	0.126	02:05
Breath Volume	1.437	02:05
Air Blank	0.000	02:05
Control Test	0.081	02:06
Air Blank	0.000	02:06
Diagnostics Check	OK	02:06

---

Data Download Date/Time: 11/20/2019 10:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BONSALL, MATTHEW, A  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: BONSALL, M  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: ARREOLA-FARFAN

DATE/TIME OF TEST: 11/05/2019 22:14:09

---

Diagnostics Check	OK	22:05
Air Blank	0.000	22:06
Control Test	0.081	22:06
Air Blank	0.000	22:06
Subject Sample # 1	0.124	22:08
Breath Volume	3.167	22:08
Air Blank	0.000	22:09
Air Blank	0.000	22:10
Subject Sample # 2	0.118	22:12
Breath Volume	3.039	22:12
Air Blank	0.000	22:13
Control Test	0.080	22:13
Air Blank	0.000	22:14
Diagnostics Check	OK	22:14

---

Data Download Date/Time: 11/20/2019 10:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BONSALL, MATTHEW, A  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: MARRERO, A  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: VARELA

DATE/TIME OF TEST: 11/06/2019 01:42:00

---

Diagnostics Check	OK	01:35
Air Blank	0.000	01:35
Control Test	0.080	01:36
Air Blank	0.000	01:36
Subject Sample # 1	0.267	01:37
Breath Volume	3.851	01:37
Air Blank	0.000	01:38
Air Blank	0.000	01:39
Subject Sample # 2	0.255	01:40
Breath Volume	3.738	01:40
Air Blank	0.000	01:41
Control Test	0.079	01:41
Air Blank	0.000	01:41
Diagnostics Check	OK	01:42

---

Data Download Date/Time: 11/20/2019 10:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHILKE,CARL,E  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: SCHILKE,CARL  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 16:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: COLON

DATE/TIME OF TEST: 11/06/2019 17:57:04

---

Diagnostics Check	OK	17:50
Air Blank	0.000	17:51
Control Test	0.080	17:51
Air Blank	0.000	17:51
Subject Sample # 1	0.107	17:52
Breath Volume	4.093	17:52
Air Blank	0.000	17:53
Air Blank	0.000	17:54
Subject Sample # 2	0.102	17:55
Breath Volume	3.710	17:55
Air Blank	0.000	17:56
Control Test	0.080	17:56
Air Blank	0.000	17:56
Diagnostics Check	OK	17:57

---

Data Download Date/Time: 11/20/2019 10:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOULD,LEONARD,  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: GOULD,LEONARD  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: GARCIA SILVA

DATE/TIME OF TEST: 11/11/2019 02:05:59

---

Diagnostics Check		OK	01:59
Air Blank		0.000	02:00
Control Test		0.080	02:00
Air Blank		0.000	02:00
Subject Sample	# 1	0.148	02:01
Breath Volume		1.554	02:01
Air Blank		0.000	02:01
Air Blank		0.000	02:03
Subject Sample	# 2	0.140	02:04
Breath Volume		2.496	02:04
Air Blank		0.000	02:05
Control Test		0.080	02:05
Air Blank		0.000	02:05
Diagnostics Check		OK	02:05

---

Data Download Date/Time: 11/20/2019 10:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BONSALL, MATTHEW, A  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: BONSALL, M  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 16:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:39

SUBJECT'S LAST NAME: SOSA

DATE/TIME OF TEST: 11/19/2019 18:06:33

---

Diagnostics Check	OK	18:00
Air Blank	0.000	18:00
Control Test	0.081	18:00
Air Blank	0.000	18:01
Subject Sample # 1	0.000	18:01
Breath Volume	1.492	18:01
Air Blank	0.000	18:02
Air Blank	0.000	18:04
Subject Sample # 2	0.000	18:05
Breath Volume	2.292	18:05
Air Blank	0.000	18:05
Control Test	0.081	18:05
Air Blank	0.000	18:06
Diagnostics Check	OK	18:06

---

Data Download Date/Time: 11/20/2019 10:29



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BONSALL, MATTHEW, A  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: BONSALL, M  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 11/20/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:55  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: MCGLEW

DATE/TIME OF TEST: 11/22/2019 23:52:45

---

Diagnostics Check		OK	23:46
Air Blank		0.000	23:46
Control Test		0.080	23:46
Air Blank		0.000	23:47
Subject Sample	# 1	0.217	23:48
Breath Volume		2.406	23:48
Air Blank		0.000	23:48
Air Blank		0.000	23:50
Subject Sample	# 2	0.200	23:51
Breath Volume		1.335	23:51
Air Blank		0.000	23:51
Control Test		0.080	23:52
Air Blank		0.000	23:52
Diagnostics Check		OK	23:52

---

Data Download Date/Time: 12/11/2019 10:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ZARRILLO,N,  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: ZARRILLO,N  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 11/20/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 18:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:07

SUBJECT'S LAST NAME: BYERS

DATE/TIME OF TEST: 11/23/2019 19:33:43

---

Diagnostics Check	OK	19:27
Air Blank	0.000	19:27
Control Test	0.080	19:28
Air Blank	0.000	19:28
Subject Sample # 1	0.124	19:29
Breath Volume	4.058	19:29
Air Blank	0.000	19:29
Air Blank	0.000	19:31
Subject Sample # 2	0.122	19:32
Breath Volume	4.812	19:32
Air Blank	0.000	19:32
Control Test	0.079	19:33
Air Blank	0.000	19:33
Diagnostics Check	OK	19:33

---

Data Download Date/Time: 12/11/2019 10:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ZARRILLO,N,  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: ZARRILLO,N  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 11/20/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: MUNIZ GAYTAN

DATE/TIME OF TEST: 11/24/2019 01:40:36

---

Diagnostics Check	OK	01:34
Air Blank	0.000	01:34
Control Test	0.081	01:35
Air Blank	0.000	01:35
Subject Sample # 1	0.134	01:36
Breath Volume	3.589	01:36
Air Blank	0.000	01:36
Air Blank	0.000	01:38
Subject Sample # 2	0.130	01:39
Breath Volume	3.652	01:39
Air Blank	0.000	01:39
Control Test	0.080	01:39
Air Blank	0.000	01:40
Diagnostics Check	OK	01:40

---

Data Download Date/Time: 12/11/2019 10:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AVENDANO,JUAN,M  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: ECKERT,KYLE  
ARRESTING OFFICER AGENCY: Sanibel PD  
LAST AGENCY INSPECTION: 11/20/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: NAGLE

DATE/TIME OF TEST: 11/26/2019 00:11:29

---

Diagnostics Check	OK	00:04
Air Blank	0.000	00:05
Control Test	0.081	00:05
Air Blank	0.000	00:06
Subject Sample # 1	0.000	00:07
Breath Volume	3.718	00:07
Air Blank	0.000	00:07
Air Blank	0.000	00:09
Subject Sample # 2	0.000	00:10
Breath Volume	3.019	00:10
Air Blank	0.000	00:10
Control Test	0.080	00:10
Air Blank	0.000	00:11
Diagnostics Check	OK	00:11

---

Data Download Date/Time: 12/11/2019 10:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CUTIE, EDWARD,  
BREATH TEST OPERATOR AGENCY: Fort Myers PD  
ARRESTING OFFICER: DOTSON, WILLIAM  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 11/20/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 16:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:57

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 11/27/2019 17:32:52

---

Diagnostics Check	OK	17:26
Air Blank	0.000	17:27
Control Test	0.080	17:27
Air Blank	0.000	17:27
Subject Sample # 1	0.000	17:28
Breath Volume	2.792	17:28
Air Blank	0.000	17:28
Air Blank	0.000	17:30
Subject Sample # 2	0.000	17:31
Breath Volume	2.109	17:31
Air Blank	0.000	17:31
Control Test	0.080	17:32
Air Blank	0.000	17:32
Diagnostics Check	OK	17:32

---

Data Download Date/Time: 12/11/2019 10:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VANPELT,J,  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: SHOEMAKER,J  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 11/20/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 18:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: TIRADO MORALES

DATE/TIME OF TEST: 11/27/2019 19:26:26

---

Diagnostics Check	OK	19:20
Air Blank	0.000	19:20
Control Test	0.081	19:20
Air Blank	0.000	19:21
Subject Sample # 1	0.000	19:21
Breath Volume	2.425	19:21
Air Blank	0.000	19:22
Air Blank	0.000	19:24
Subject Sample # 2	0.000	19:25
Breath Volume	3.601	19:25
Air Blank	0.000	19:25
Control Test	0.081	19:25
Air Blank	0.000	19:26
Diagnostics Check	OK	19:26

---

Data Download Date/Time: 12/11/2019 10:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STROBEL, RAYMOND, R  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: STROBEL, RAYMOND  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 11/20/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 14:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:24

SUBJECT'S LAST NAME: KONLEY

DATE/TIME OF TEST: 11/28/2019 14:55:00

---

Diagnostics Check	OK	14:52	
Air Blank	0.000	14:52	
Control Test	0.081	14:52	
Air Blank	0.000	14:53	
Subject Sample	# 1	14:53	Subject Test Refused
Breath Volume	0.000	14:53	
Air Blank	0.000	14:54	
Control Test	0.080	14:54	
Air Blank	0.000	14:54	
Diagnostics Check	OK	14:55	

---

Data Download Date/Time: 12/11/2019 10:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BONSALL, MATTHEW, A  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: BONSALL, M  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 11/20/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: BARNETT

DATE/TIME OF TEST: 11/29/2019 03:14:00

---

Diagnostics Check		OK	03:07
Air Blank		0.000	03:08
Control Test		0.081	03:08
Air Blank		0.000	03:09
Subject Sample	# 1	0.184	03:09
Breath Volume		1.894	03:09
Air Blank		0.000	03:10
Air Blank		0.000	03:12
Subject Sample	# 2	0.172	03:12
Breath Volume		1.261	03:12
Air Blank		0.000	03:13
Control Test		0.080	03:13
Air Blank		0.000	03:13
Diagnostics Check		OK	03:14

---

Data Download Date/Time: 12/11/2019 10:19



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FARRIS, MICHAEL, T  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: FARRIS, MICHAEL  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 11/20/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:54

SUBJECT'S LAST NAME: BRINDLEY

DATE/TIME OF TEST: 11/29/2019 20:21:59

---

Diagnostics Check	OK	20:15
Air Blank	0.000	20:15
Control Test	0.080	20:16
Air Blank	0.000	20:16
Subject Sample # 1	0.073	20:17
Breath Volume	1.789	20:17
Air Blank	0.000	20:17
Air Blank	0.000	20:19
Subject Sample # 2	0.081	20:20
Breath Volume	1.425	20:20
Air Blank	0.000	20:21
Control Test	0.081	20:21
Air Blank	0.000	20:21
Diagnostics Check	OK	20:21

---

Data Download Date/Time: 12/11/2019 10:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRAY,DANNY,W  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: AUBRY,MICHAEL  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: VENEIGH

DATE/TIME OF TEST: 11/02/2019 02:30:14

---

Diagnostics Check	OK	02:23
Air Blank	0.000	02:23
Control Test	0.080	02:24
Air Blank	0.000	02:24
Subject Sample # 1	0.181	02:25
Breath Volume	3.546	02:25
Air Blank	0.000	02:26
Air Blank	0.000	02:27
Subject Sample # 2	0.178	02:28
Breath Volume	4.179	02:28
Air Blank	0.000	02:29
Control Test	0.079	02:29
Air Blank	0.000	02:30
Diagnostics Check	OK	02:30

---

Data Download Date/Time: 11/20/2019 10:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ENGH, WILLIAM, C  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: LOETHEN, K  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:02  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: HEROUX

DATE/TIME OF TEST: 11/02/2019 22:40:45

---

Diagnostics Check	OK	22:34
Air Blank	0.000	22:34
Control Test	0.080	22:34
Air Blank	0.000	22:35
Subject Sample # 1	0.128	22:36
Breath Volume	2.601	22:36
Air Blank	0.000	22:36
Air Blank	0.000	22:38
Subject Sample # 2	0.129	22:39
Breath Volume	1.761	22:39
Air Blank	0.000	22:39
Control Test	0.079	22:40
Air Blank	0.000	22:40
Diagnostics Check	OK	22:40

---

Data Download Date/Time: 11/20/2019 10:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROSS,CHEYANNE,K  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: OHERN,ANDREW  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: AMEZQUITA CHIRAS

DATE/TIME OF TEST: 11/03/2019 01:42:20

---

Diagnostics Check	OK	01:35
Air Blank	0.000	01:36
Control Test	0.080	01:36
Air Blank	0.000	01:37
Subject Sample # 1	0.112	01:37
Breath Volume	1.851	01:37
Air Blank	0.000	01:38
Air Blank	0.000	01:40
Subject Sample # 2	0.113	01:40
Breath Volume	1.589	01:40
Air Blank	0.000	01:41
Control Test	0.080	01:41
Air Blank	0.000	01:42
Diagnostics Check	OK	01:42

---

Data Download Date/Time: 11/20/2019 10:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FILLAK, MICHAEL, O  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: PEHNS, TIMOTHY  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: DUARTE DUENAS

DATE/TIME OF TEST: 11/04/2019 03:56:32

---

Diagnostics Check	OK	03:48
Air Blank	0.000	03:49
Control Test	0.080	03:49
Air Blank	0.000	03:50
Subject Sample # 1	0.145	03:50
Breath Volume	1.496	03:50
Air Blank	0.000	03:51
Air Blank	0.000	03:53
Subject Sample # 2	0.144	03:54
Breath Volume	1.753	03:54
Air Blank	0.000	03:55
Control Test	0.079	03:55
Air Blank	0.000	03:56
Diagnostics Check	OK	03:56

---

Data Download Date/Time: 11/20/2019 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYERS,JUSTIN,L  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: MYERS,JUSTIN  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: ROOT

DATE/TIME OF TEST: 11/06/2019 00:28:17

---

Diagnostics Check	OK	00:20
Air Blank	0.000	00:20
Control Test	0.080	00:21
Air Blank	0.000	00:21
Subject Sample # 1	0.092	00:22
Breath Volume	1.257	00:22
Air Blank	0.000	00:23
Air Blank	0.000	00:25
Subject Sample # 2	0.090	00:26
Breath Volume	1.738	00:26
Air Blank	0.000	00:27
Control Test	0.080	00:27
Air Blank	0.000	00:28
Diagnostics Check	OK	00:28

---

Data Download Date/Time: 11/20/2019 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ENGH,W,C  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: VENSKI,J  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:09  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:28

SUBJECT'S LAST NAME: AVERY

DATE/TIME OF TEST: 11/06/2019 21:54:34

---

Diagnostics Check		OK	21:48
Air Blank		0.000	21:48
Control Test		0.080	21:49
Air Blank		0.000	21:49
Subject Sample	# 1	0.014	21:50
Breath Volume		1.468	21:50
Air Blank		0.000	21:50
Air Blank		0.000	21:52
Subject Sample	# 2	0.014	21:53
Breath Volume		1.105	21:53
Air Blank		0.000	21:53
Control Test		0.081	21:53
Air Blank		0.000	21:54
Diagnostics Check		OK	21:54

---

Data Download Date/Time: 11/20/2019 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTINI,RYAN,H  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: MARTINI,RYAN  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:24

SUBJECT'S LAST NAME: MAREK

DATE/TIME OF TEST: 11/07/2019 21:54:21

---

Diagnostics Check	OK	21:48
Air Blank	0.000	21:48
Control Test	0.081	21:49
Air Blank	0.000	21:49
Subject Sample # 1	0.000	21:50
Breath Volume	4.285	21:50
Air Blank	0.000	21:50
Air Blank	0.000	21:52
Subject Sample # 2	0.000	21:52
Breath Volume	2.613	21:52
Air Blank	0.000	21:53
Control Test	0.080	21:53
Air Blank	0.000	21:54
Diagnostics Check	OK	21:54

---

Data Download Date/Time: 11/20/2019 10:37



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARKLEY,JEFFREY,G  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: HARRISON,WILLIE  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: KEPHART

DATE/TIME OF TEST: 11/07/2019 23:12:48

---

Diagnostics Check		OK	23:02	
Air Blank		0.000	23:03	
Control Test		0.080	23:03	
Air Blank		0.000	23:04	
Subject Sample	# 1	0.187	23:07	Volume Not Met
Breath Volume		0.066	23:07	
Air Blank		0.000	23:08	
Air Blank		0.000	23:09	
Subject Sample	# 2		23:11	Subject Test Refused
Breath Volume		0.050	23:11	
Air Blank		0.000	23:11	
Control Test		0.078	23:12	
Air Blank		0.000	23:12	
Diagnostics Check		OK	23:12	

---

Data Download Date/Time: 11/20/2019 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARKLEY,JEFFREY,G  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: HARRISON,WILLIE  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: KEPHART

DATE/TIME OF TEST: 11/08/2019 00:14:34

---

Diagnostics Check		OK	00:05	
Air Blank		0.000	00:05	
Control Test		0.079	00:05	
Air Blank		0.000	00:06	
Subject Sample	# 1	0.200	00:09	Volume Not Met
Breath Volume		0.015	00:09	
Air Blank		0.000	00:10	
Air Blank		0.000	00:11	
Subject Sample	# 2		00:13	Subject Test Refused
Breath Volume		0.000	00:13	
Air Blank		0.000	00:13	
Control Test		0.077	00:13	
Air Blank		0.000	00:14	
Diagnostics Check		OK	00:14	

---

Data Download Date/Time: 11/20/2019 10:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYERS,JUSTIN,L  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: MYERS,JUSTIN  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:57  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: CUCCHIARA

DATE/TIME OF TEST: 11/09/2019 03:54:28

---

Diagnostics Check	OK	03:47
Air Blank	0.000	03:48
Control Test	0.080	03:48
Air Blank	0.000	03:49
Subject Sample # 1	0.027	03:49
Breath Volume	1.687	03:49
Air Blank	0.000	03:50
Air Blank	0.000	03:52
Subject Sample # 2	0.026	03:52
Breath Volume	1.667	03:52
Air Blank	0.000	03:53
Control Test	0.080	03:53
Air Blank	0.000	03:54
Diagnostics Check	OK	03:54

---

Data Download Date/Time: 11/20/2019 10:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYERS,JUSTIN,L  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: HARRISON,WILLIE  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 06:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:57

SUBJECT'S LAST NAME: DA SILVA

DATE/TIME OF TEST: 11/09/2019 07:26:25

---

Diagnostics Check	OK	07:19
Air Blank	0.000	07:19
Control Test	0.080	07:19
Air Blank	0.000	07:20
Subject Sample # 1	0.134	07:21
Breath Volume	1.554	07:21
Air Blank	0.000	07:22
Air Blank	0.000	07:24
Subject Sample # 2	0.152	07:24
Breath Volume	3.582	07:24
Air Blank	0.000	07:25
Control Test	0.080	07:25
Air Blank	0.000	07:26
Diagnostics Check	OK	07:26

---

Data Download Date/Time: 11/20/2019 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HERNANDEZ,ELDA,I  
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO  
ARRESTING OFFICER: MORERA,DELIANYS  
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO  
LAST AGENCY INSPECTION: 10/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: TORRESALVARADO

DATE/TIME OF TEST: 11/03/2019 01:06:58

---

Diagnostics Check	OK	01:00
Air Blank	0.000	01:00
Control Test	0.079	01:01
Air Blank	0.000	01:01
Subject Sample # 1	0.189	01:02
Breath Volume	2.515	01:02
Air Blank	0.000	01:03
Air Blank	0.000	01:04
Subject Sample # 2	0.189	01:05
Breath Volume	3.011	01:05
Air Blank	0.000	01:06
Control Test	0.078	01:06
Air Blank	0.000	01:06
Diagnostics Check	OK	01:06

---

Data Download Date/Time: 11/04/2019 17:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HERNANDEZ,ELDA,I  
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO  
ARRESTING OFFICER: MILLER,KURT  
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO  
LAST AGENCY INSPECTION: 10/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 11/03/2019 02:34:53

---

Diagnostics Check	OK	02:28
Air Blank	0.000	02:28
Control Test	0.079	02:29
Air Blank	0.000	02:29
Subject Sample # 1	0.147	02:30
Breath Volume	3.281	02:30
Air Blank	0.000	02:30
Air Blank	0.000	02:32
Subject Sample # 2	0.143	02:33
Breath Volume	2.210	02:33
Air Blank	0.000	02:33
Control Test	0.079	02:34
Air Blank	0.000	02:34
Diagnostics Check	OK	02:34

---

Data Download Date/Time: 11/04/2019 17:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARMON,C,A  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: VASQUEZ,D  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 11/01/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: JOSEPH

DATE/TIME OF TEST: 11/09/2019 21:27:00

---

Diagnostics Check		OK	21:15	
Air Blank		0.000	21:15	
Control Test		0.081	21:15	
Air Blank		0.000	21:16	
Subject Sample	# 1	0.147	21:19	Volume Not Met
Breath Volume		0.460	21:19	
Air Blank		0.000	21:20	
Air Blank		0.000	21:21	
Subject Sample	# 2	0.278	21:22	
Breath Volume		3.945	21:22	
Air Blank		0.000	21:23	
Air Blank		0.000	21:25	
Subject Sample	# 3	0.277	21:25	
Breath Volume		4.144	21:25	
Air Blank		0.000	21:26	
Control Test		0.080	21:26	
Air Blank		0.000	21:26	
Diagnostics Check		OK	21:27	

---

Data Download Date/Time: 12/06/2019 11:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYKHASKIV,M,  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: VASQUEZ,D  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 11/01/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: NOEL

DATE/TIME OF TEST: 11/23/2019 00:52:53

---

Diagnostics Check	OK	00:44
Air Blank	0.000	00:45
Control Test	0.081	00:45
Air Blank	0.000	00:45
Subject Sample # 1	0.159	00:47
Breath Volume	1.785	00:47
Air Blank	0.000	00:48
Air Blank	0.000	00:50
Subject Sample # 2	0.150	00:51
Breath Volume	1.457	00:51
Air Blank	0.000	00:51
Control Test	0.080	00:52
Air Blank	0.000	00:52
Diagnostics Check	OK	00:52

---

Data Download Date/Time: 12/06/2019 11:58



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LOPEZ,JOSE,D  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: LOPEZ,JOSE  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 11/01/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: LOPEZLOPEZ

DATE/TIME OF TEST: 11/24/2019 04:14:41

---

Diagnostics Check		OK	04:08
Air Blank		0.000	04:08
Control Test		0.081	04:08
Air Blank		0.000	04:09
Subject Sample	# 1	0.226	04:10
Breath Volume		2.539	04:10
Air Blank		0.000	04:10
Air Blank		0.000	04:12
Subject Sample	# 2	0.227	04:13
Breath Volume		2.250	04:13
Air Blank		0.000	04:13
Control Test		0.081	04:14
Air Blank		0.000	04:14
Diagnostics Check		OK	04:14

---

Data Download Date/Time: 12/06/2019 11:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000949

INSTRUMENT REGISTERED TO: TALLAHASSEE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KELLER,DAVID,  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: KELLER,DAVID  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 10/30/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A1  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: ARTHURS

DATE/TIME OF TEST: 11/07/2019 23:45:47

---

Diagnostics Check	OK	23:42	
Air Blank	0.000	23:43	
Control Test	0.080	23:43	
Air Blank	0.000	23:43	
Subject Sample # 1	0.238	23:44	
Breath Volume	2.011	23:44	
Air Blank		23:44	RFI Detect
Air Blank	0.000	23:44	
Control Test	0.080	23:45	
Air Blank	0.000	23:45	
Diagnostics Check	OK	23:45	

---

Data Download Date/Time: 11/25/2019 09:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000950

INSTRUMENT REGISTERED TO: HAVANA PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOSCHETTO, ANTHONY, C  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: MOSCHETTO, ANTHONY  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 10/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29017080A3  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 12:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 12:40

SUBJECT'S LAST NAME: HUBER

DATE/TIME OF TEST: 11/04/2019 13:45:56

---

Diagnostics Check	OK	13:38
Air Blank	0.000	13:39
Control Test	0.075	13:39
Air Blank	0.000	13:40
Subject Sample # 1	0.147	13:41
Breath Volume	3.589	13:41
Air Blank	0.000	13:41
Air Blank	0.000	13:43
Subject Sample # 2	0.146	13:44
Breath Volume	3.289	13:44
Air Blank	0.000	13:45
Control Test	0.075	13:45
Air Blank	0.000	13:45
Diagnostics Check	OK	13:45

---

Data Download Date/Time: 11/05/2019 11:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: MORALES,JOSE  
ARRESTING OFFICER AGENCY: Central FL Univ PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: JORDAN

DATE/TIME OF TEST: 11/01/2019 02:30:40

---

Diagnostics Check	OK	02:24
Air Blank	0.000	02:25
Control Test	0.081	02:25
Air Blank	0.000	02:25
Subject Sample # 1	0.139	02:26
Breath Volume	2.328	02:26
Air Blank	0.000	02:26
Air Blank	0.000	02:28
Subject Sample # 2	0.142	02:29
Breath Volume	2.605	02:29
Air Blank	0.000	02:29
Control Test	0.081	02:30
Air Blank	0.000	02:30
Diagnostics Check	OK	02:30

---

Data Download Date/Time: 11/07/2019 15:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TIAMSON,DAWN,P  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: MOREL,ABIGAIL  
ARRESTING OFFICER AGENCY: Maitland PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: DELANCEY

DATE/TIME OF TEST: 11/01/2019 03:15:46

---

Diagnostics Check	OK	03:09
Air Blank	0.000	03:09
Control Test	0.082	03:10
Air Blank	0.000	03:10
Subject Sample # 1	0.100	03:11
Breath Volume	2.054	03:11
Air Blank	0.000	03:11
Air Blank	0.000	03:13
Subject Sample # 2	0.104	03:14
Breath Volume	2.890	03:14
Air Blank	0.000	03:14
Control Test	0.082	03:15
Air Blank	0.000	03:15
Diagnostics Check	OK	03:15

---

Data Download Date/Time: 11/07/2019 15:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: JOHNS,MARK  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:28  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: HERBAS GUTIERREZ

DATE/TIME OF TEST: 11/01/2019 04:23:06

---

Diagnostics Check		OK	04:11	
Air Blank		0.000	04:11	
Control Test		0.082	04:12	
Air Blank		0.000	04:12	
Subject Sample	# 1	0.193	04:15	Volume Not Met
Breath Volume		0.164	04:15	
Air Blank		0.000	04:16	
Air Blank		0.000	04:18	
Subject Sample	# 2	0.178	04:21	Volume Not Met
Breath Volume		0.410	04:21	
Air Blank		0.000	04:22	
Control Test		0.080	04:22	
Air Blank		0.000	04:23	
Diagnostics Check		OK	04:23	

---

Data Download Date/Time: 11/07/2019 15:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: WISE,IAN  
ARRESTING OFFICER AGENCY: OTHER Local  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:36

SUBJECT'S LAST NAME: OTERO

DATE/TIME OF TEST: 11/01/2019 21:05:04

---

Diagnostics Check	OK	20:59
Air Blank	0.000	20:59
Control Test	0.081	20:59
Air Blank	0.000	21:00
Subject Sample # 1	0.000	21:00
Breath Volume	4.753	21:00
Air Blank	0.000	21:01
Air Blank	0.000	21:03
Subject Sample # 2	0.000	21:03
Breath Volume	4.703	21:03
Air Blank	0.000	21:04
Control Test	0.081	21:04
Air Blank	0.000	21:04
Diagnostics Check	OK	21:05

---

Data Download Date/Time: 11/07/2019 15:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: GILZEN,MARCUS  
ARRESTING OFFICER AGENCY: Apopka PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: BELTRAN ROMAN

DATE/TIME OF TEST: 11/02/2019 02:43:40

---

Diagnostics Check	OK	02:37
Air Blank	0.000	02:37
Control Test	0.080	02:38
Air Blank	0.000	02:38
Subject Sample # 1	0.111	02:39
Breath Volume	2.402	02:39
Air Blank	0.000	02:39
Air Blank	0.000	02:41
Subject Sample # 2	0.105	02:42
Breath Volume	1.441	02:42
Air Blank	0.000	02:42
Control Test	0.081	02:43
Air Blank	0.000	02:43
Diagnostics Check	OK	02:43

---

Data Download Date/Time: 11/07/2019 15:33



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DENSON,TRAVIS,W  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: MARTINEZ,ERNESTO  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: OCHOA GARCIA

DATE/TIME OF TEST: 11/02/2019 03:23:40

---

Diagnostics Check		OK	03:09	
Air Blank		0.000	03:10	
Control Test		0.081	03:10	
Air Blank		0.000	03:11	
Subject Sample	# 1	0.141	03:14	Volume Not Met
Breath Volume		0.187	03:14	
Air Blank		0.000	03:14	
Air Blank		0.000	03:16	
Subject Sample	# 2	0.180	03:19	
Breath Volume		2.101	03:19	
Air Blank		0.000	03:20	
Air Blank		0.000	03:21	
Subject Sample	# 3	0.175	03:22	
Breath Volume		1.367	03:22	
Air Blank		0.000	03:22	
Control Test		0.078	03:23	
Air Blank		0.000	03:23	
Diagnostics Check		OK	03:23	

---

Data Download Date/Time: 11/07/2019 15:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: WISE,IAN  
ARRESTING OFFICER AGENCY: OTHER Local  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: RODRIGUEZ JR

DATE/TIME OF TEST: 11/02/2019 03:57:18

---

Diagnostics Check		OK	03:50
Air Blank		0.000	03:50
Control Test		0.081	03:51
Air Blank		0.000	03:51
Subject Sample	# 1	0.093	03:52
Breath Volume		1.753	03:52
Air Blank		0.000	03:53
Air Blank		0.000	03:54
Subject Sample	# 2	0.089	03:55
Breath Volume		1.335	03:55
Air Blank		0.000	03:56
Control Test		0.080	03:56
Air Blank		0.000	03:57
Diagnostics Check		OK	03:57

---

Data Download Date/Time: 11/07/2019 15:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DENSON, TRAVIS, W  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: JUNCO, JONATHAN  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: CHUNILALL

DATE/TIME OF TEST: 11/02/2019 04:08:56

---

Diagnostics Check	OK	04:03
Air Blank	0.000	04:03
Control Test	0.081	04:03
Air Blank	0.000	04:04
Subject Sample # 1	0.049	04:04
Breath Volume	1.785	04:04
Air Blank	0.000	04:05
Air Blank	0.000	04:07
Subject Sample # 2	0.047	04:07
Breath Volume	2.113	04:07
Air Blank	0.000	04:08
Control Test	0.081	04:08
Air Blank	0.000	04:08
Diagnostics Check	OK	04:08

---

Data Download Date/Time: 11/07/2019 15:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: WISE,IAN  
ARRESTING OFFICER AGENCY: OTHER Local  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: SAENZ

DATE/TIME OF TEST: 11/03/2019 04:29:04

---

Diagnostics Check	OK	04:22
Air Blank	0.000	04:22
Control Test	0.081	04:23
Air Blank	0.000	04:23
Subject Sample # 1	0.217	04:24
Breath Volume	1.617	04:24
Air Blank	0.000	04:25
Air Blank	0.000	04:27
Subject Sample # 2	0.228	04:27
Breath Volume	1.910	04:27
Air Blank	0.000	04:28
Control Test	0.081	04:28
Air Blank	0.000	04:28
Diagnostics Check	OK	04:29

---

Data Download Date/Time: 11/07/2019 15:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NERVAIS,TABITHA,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: WISE,IAN  
ARRESTING OFFICER AGENCY: OTHER Local  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:13  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: RODRIGUEZDIAZ

DATE/TIME OF TEST: 11/04/2019 04:00:56

---

Diagnostics Check	OK	03:51
Air Blank	0.000	03:52
Control Test	0.081	03:52
Air Blank	0.000	03:53
Subject Sample # 1	0.114	03:54
Breath Volume	1.187	03:54
Air Blank	0.000	03:55
Air Blank	0.000	03:57
Subject Sample # 2	0.112	03:59
Breath Volume	2.492	03:59
Air Blank	0.000	04:00
Control Test	0.082	04:00
Air Blank	0.000	04:00
Diagnostics Check	OK	04:00

---

Data Download Date/Time: 11/07/2019 15:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NERVAIS,TABITHA,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: LAWRENCE,NICK  
ARRESTING OFFICER AGENCY: Maitland PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:27  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 11/04/2019 23:25:51

---

Diagnostics Check	OK	23:19
Air Blank	0.000	23:20
Control Test	0.082	23:20
Air Blank	0.000	23:21
Subject Sample # 1	0.000	23:21
Breath Volume	2.363	23:21
Air Blank	0.000	23:22
Air Blank	0.000	23:24
Subject Sample # 2	0.000	23:24
Breath Volume	2.527	23:24
Air Blank	0.000	23:24
Control Test	0.082	23:25
Air Blank	0.000	23:25
Diagnostics Check	OK	23:25

---

Data Download Date/Time: 11/07/2019 15:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TIAMSON,DAWN,P  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: BRITT,JOHN  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: SWANSON

DATE/TIME OF TEST: 11/04/2019 23:54:37

---

Diagnostics Check	OK	23:43	
Air Blank	0.000	23:43	
Control Test	0.081	23:43	
Air Blank	0.000	23:44	
Subject Sample	# 1 0.190	23:47	Volume Not Met
Breath Volume	0.343	23:47	
Air Blank	0.000	23:48	
Air Blank	0.000	23:49	
Subject Sample	# 2 0.164	23:53	Volume Not Met
Breath Volume	0.167	23:53	
Air Blank	0.000	23:53	
Control Test	0.078	23:54	
Air Blank	0.000	23:54	
Diagnostics Check	OK	23:54	

---

Data Download Date/Time: 11/07/2019 15:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILKINS,CONNOR,R  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: KNOWLES,ZOE  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 11/05/2019 00:43:49

---

Diagnostics Check	OK	00:37
Air Blank	0.000	00:37
Control Test	0.081	00:38
Air Blank	0.000	00:38
Subject Sample # 1	0.000	00:39
Breath Volume	2.191	00:39
Air Blank	0.000	00:40
Air Blank	0.000	00:42
Subject Sample # 2	0.000	00:42
Breath Volume	3.316	00:42
Air Blank	0.000	00:42
Control Test	0.081	00:43
Air Blank	0.000	00:43
Diagnostics Check	OK	00:43

---

Data Download Date/Time: 11/07/2019 15:33



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NERVAIS,TABITHA,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: MORALES,JOSE  
ARRESTING OFFICER AGENCY: Central FL Univ PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 11/05/2019 03:47:24

---

Diagnostics Check	OK	03:40
Air Blank	0.000	03:40
Control Test	0.082	03:41
Air Blank	0.000	03:41
Subject Sample # 1	0.118	03:42
Breath Volume	1.531	03:42
Air Blank	0.000	03:43
Air Blank	0.000	03:45
Subject Sample # 2	0.114	03:45
Breath Volume	1.621	03:45
Air Blank	0.000	03:46
Control Test	0.081	03:46
Air Blank	0.000	03:47
Diagnostics Check	OK	03:47

---

Data Download Date/Time: 11/07/2019 15:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TIAMSON,DAWN,P  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: FERGUSON,STEVEN  
ARRESTING OFFICER AGENCY: Winter Park PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:11  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: WARNER

DATE/TIME OF TEST: 11/07/2019 05:12:29

---

Diagnostics Check	OK	05:05
Air Blank	0.000	05:05
Control Test	0.081	05:06
Air Blank	0.000	05:06
Subject Sample # 1	0.196	05:07
Breath Volume	2.332	05:07
Air Blank	0.000	05:07
Air Blank	0.000	05:09
Subject Sample # 2	0.198	05:10
Breath Volume	1.605	05:10
Air Blank	0.000	05:11
Control Test	0.081	05:11
Air Blank	0.000	05:12
Diagnostics Check	OK	05:12

---

Data Download Date/Time: 11/07/2019 15:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: PETULLA,KYLE  
ARRESTING OFFICER AGENCY: Apopka PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:42  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 11/08/2019 04:02:18

---

Diagnostics Check	OK	03:55
Air Blank	0.000	03:56
Control Test	0.082	03:56
Air Blank	0.000	03:56
Subject Sample # 1	0.182	03:57
Breath Volume	1.316	03:57
Air Blank	0.000	03:58
Air Blank	0.000	04:00
Subject Sample # 2	0.188	04:00
Breath Volume	1.371	04:00
Air Blank	0.000	04:01
Control Test	0.081	04:01
Air Blank	0.000	04:02
Diagnostics Check	OK	04:02

---

Data Download Date/Time: 11/11/2019 13:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: LAL,RAJIV  
ARRESTING OFFICER AGENCY: OTHER Local  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: DELVARICE

DATE/TIME OF TEST: 11/09/2019 04:21:18

---

Diagnostics Check	OK	04:14
Air Blank	0.000	04:15
Control Test	0.082	04:15
Air Blank	0.000	04:15
Subject Sample # 1	0.206	04:16
Breath Volume	1.156	04:16
Air Blank	0.000	04:16
Air Blank	0.000	04:18
Subject Sample # 2	0.195	04:19
Breath Volume	1.195	04:19
Air Blank	0.000	04:20
Control Test	0.081	04:20
Air Blank	0.000	04:21
Diagnostics Check	OK	04:21

---

Data Download Date/Time: 11/11/2019 13:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: WILK,JULIE  
ARRESTING OFFICER AGENCY: Central FL Univ PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 11/09/2019 05:03:03

---

Diagnostics Check	OK	04:56
Air Blank	0.000	04:57
Control Test	0.082	04:57
Air Blank	0.000	04:58
Subject Sample # 1	0.000	04:58
Breath Volume	3.921	04:58
Air Blank	0.000	04:59
Air Blank	0.000	05:01
Subject Sample # 2	0.000	05:01
Breath Volume	3.875	05:01
Air Blank	0.000	05:02
Control Test	0.081	05:02
Air Blank	0.000	05:02
Diagnostics Check	OK	05:03

---

Data Download Date/Time: 11/11/2019 13:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: WADEMAN,CHRISTOPHER  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 08:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:05

SUBJECT'S LAST NAME: ETHIER

DATE/TIME OF TEST: 11/09/2019 09:36:26

---

Diagnostics Check		OK	09:30
Air Blank		0.000	09:30
Control Test		0.082	09:31
Air Blank		0.000	09:31
Subject Sample	# 1	0.249	09:32
Breath Volume		2.628	09:32
Air Blank		0.000	09:32
Air Blank		0.000	09:34
Subject Sample	# 2	0.245	09:34
Breath Volume		2.363	09:34
Air Blank		0.000	09:35
Control Test		0.080	09:35
Air Blank		0.000	09:36
Diagnostics Check		OK	09:36

---

Data Download Date/Time: 11/11/2019 13:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILKINS,CONNOR,R  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: MCCONACHIE,PATRICK  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: HOLLRAH

DATE/TIME OF TEST: 11/09/2019 22:33:11

---

Diagnostics Check	OK	22:26
Air Blank	0.000	22:26
Control Test	0.082	22:27
Air Blank	0.000	22:27
Subject Sample # 1	0.144	22:28
Breath Volume	1.500	22:28
Air Blank	0.000	22:28
Air Blank	0.000	22:30
Subject Sample # 2	0.162	22:31
Breath Volume	3.613	22:31
Air Blank	0.000	22:32
Control Test	0.081	22:32
Air Blank	0.000	22:33
Diagnostics Check	OK	22:33

---

Data Download Date/Time: 11/11/2019 13:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: SPEER,ANDREW  
ARRESTING OFFICER AGENCY: Maitland PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 11/10/2019 02:23:37

---

Diagnostics Check	OK	02:16
Air Blank	0.000	02:17
Control Test	0.082	02:17
Air Blank	0.000	02:18
Subject Sample # 1	0.159	02:19
Breath Volume	3.535	02:19
Air Blank	0.000	02:19
Air Blank	0.000	02:21
Subject Sample # 2	0.144	02:22
Breath Volume	1.996	02:22
Air Blank	0.000	02:22
Control Test	0.080	02:23
Air Blank	0.000	02:23
Diagnostics Check	OK	02:23

---

Data Download Date/Time: 11/11/2019 13:44



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: PAULINO,JOSE  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: SOTO

DATE/TIME OF TEST: 11/10/2019 04:23:12

---

Diagnostics Check	OK	04:16
Air Blank	0.000	04:16
Control Test	0.082	04:17
Air Blank	0.000	04:17
Subject Sample # 1	0.178	04:18
Breath Volume	1.527	04:18
Air Blank	0.000	04:19
Air Blank	0.000	04:21
Subject Sample # 2	0.168	04:21
Breath Volume	1.531	04:21
Air Blank	0.000	04:22
Control Test	0.080	04:22
Air Blank	0.000	04:23
Diagnostics Check	OK	04:23

---

Data Download Date/Time: 11/11/2019 13:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILKINS,CONNOR,R  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: RAMIREZ,MATTHEW  
ARRESTING OFFICER AGENCY: Ocoee PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: MURILLOVARELA

DATE/TIME OF TEST: 11/10/2019 05:17:42

---

Diagnostics Check		OK	05:05	
Air Blank		0.000	05:06	
Control Test		0.082	05:06	
Air Blank		0.000	05:06	
Subject Sample	# 1	0.121	05:10	Volume Not Met
Breath Volume		0.453	05:10	
Air Blank		0.000	05:10	
Air Blank		0.000	05:12	
Subject Sample	# 2	0.136	05:13	
Breath Volume		1.304	05:13	
Air Blank		0.000	05:13	
Air Blank		0.000	05:15	
Subject Sample	# 3	0.136	05:16	
Breath Volume		1.738	05:16	
Air Blank		0.000	05:16	
Control Test		0.081	05:17	
Air Blank		0.000	05:17	
Diagnostics Check		OK	05:17	

---

Data Download Date/Time: 11/11/2019 13:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DENSON, TRAVIS, W  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: REDWAN, ALFREDO  
ARRESTING OFFICER AGENCY: FHP Troop K  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: GABRIEL

DATE/TIME OF TEST: 11/10/2019 19:56:33

---

Diagnostics Check	OK	19:50
Air Blank	0.000	19:50
Control Test	0.081	19:51
Air Blank	0.000	19:51
Subject Sample # 1	0.162	19:52
Breath Volume	1.929	19:52
Air Blank	0.000	19:52
Air Blank	0.000	19:54
Subject Sample # 2	0.160	19:55
Breath Volume	1.722	19:55
Air Blank	0.000	19:55
Control Test	0.081	19:55
Air Blank	0.000	19:56
Diagnostics Check	OK	19:56

---

Data Download Date/Time: 11/11/2019 13:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DENSON,TRAVIS,W  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: JOHNS,MARK  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:01  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: BERROS TORRES

DATE/TIME OF TEST: 11/11/2019 00:50:02

---

Diagnostics Check	OK	00:46	
Air Blank	0.000	00:46	
Control Test	0.082	00:47	
Air Blank	0.000	00:47	
Subject Sample	# 1	00:48	Subject Test Refused
Breath Volume	0.000	00:48	
Air Blank	0.000	00:49	
Control Test	0.082	00:49	
Air Blank	0.000	00:49	
Diagnostics Check	OK	00:50	

---

Data Download Date/Time: 11/11/2019 13:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVIS,COURTNEY,L  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: SPEER,ANDREW  
ARRESTING OFFICER AGENCY: Maitland PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:56  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: GENTRY

DATE/TIME OF TEST: 11/11/2019 02:08:41

---

Diagnostics Check		OK	02:01
Air Blank		0.000	02:02
Control Test		0.082	02:02
Air Blank		0.000	02:03
Subject Sample	# 1	0.137	02:04
Breath Volume		2.667	02:04
Air Blank		0.000	02:04
Air Blank		0.000	02:06
Subject Sample	# 2	0.132	02:07
Breath Volume		1.980	02:07
Air Blank		0.000	02:07
Control Test		0.080	02:08
Air Blank		0.000	02:08
Diagnostics Check		OK	02:08

---

Data Download Date/Time: 11/11/2019 13:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVIS,COURTNEY,L  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: BINET,MARLON  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: SALAS RIVERA

DATE/TIME OF TEST: 11/11/2019 03:10:29

---

Diagnostics Check		OK	02:57	
Air Blank		0.000	02:58	
Control Test		0.082	02:58	
Air Blank		0.000	02:58	
Subject Sample	# 1	0.207	03:02	Volume Not Met
Breath Volume		0.117	03:02	
Air Blank		0.000	03:02	
Air Blank		0.000	03:04	
Subject Sample	# 2	0.258	03:05	
Breath Volume		1.527	03:05	
Air Blank		0.000	03:06	
Air Blank		0.000	03:08	
Subject Sample	# 3	0.260	03:08	
Breath Volume		2.054	03:08	
Air Blank		0.000	03:09	
Control Test		0.080	03:09	
Air Blank		0.000	03:10	
Diagnostics Check		OK	03:10	

---

Data Download Date/Time: 11/11/2019 13:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TIAMSON,DAWN,P  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: BRAZEE,COLE  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 11/11/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: GARCIALOPEZ

DATE/TIME OF TEST: 11/14/2019 02:52:09

---

Diagnostics Check	OK	02:40	
Air Blank	0.000	02:41	
Control Test	0.081	02:41	
Air Blank	0.000	02:42	
Subject Sample	# 1	02:45	No Sample Provided
Breath Volume	0.000	02:45	
Air Blank	0.000	02:45	
Air Blank	0.000	02:47	
Subject Sample	# 2	02:50	No Sample Provided
Breath Volume	0.000	02:50	
Air Blank	0.000	02:51	
Control Test	0.081	02:51	
Air Blank	0.000	02:52	
Diagnostics Check	OK	02:52	

---

Data Download Date/Time: 11/21/2019 16:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: KAREEM,LAITH  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 11/11/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 11/23/2019 21:53:55

---

Diagnostics Check	OK	21:45
Air Blank	0.000	21:46
Control Test	0.082	21:46
Air Blank	0.000	21:47
Subject Sample # 1	0.021	21:48
Breath Volume	1.109	21:48
Air Blank	0.000	21:49
Air Blank	0.000	21:51
Subject Sample # 2	0.021	21:52
Breath Volume	1.886	21:52
Air Blank	0.000	21:53
Control Test	0.081	21:53
Air Blank	0.000	21:53
Diagnostics Check	OK	21:53

---

Data Download Date/Time: 11/28/2019 15:20



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARDEN,WILLIAM,L  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: HANSTINE,AUSTIN  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: HERNANDEZ MARTINEZ

DATE/TIME OF TEST: 11/01/2019 00:18:26

---

Diagnostics Check	OK	00:11
Air Blank	0.000	00:11
Control Test	0.079	00:12
Air Blank	0.000	00:12
Subject Sample # 1	0.090	00:14
Breath Volume	1.761	00:14
Air Blank	0.000	00:14
Air Blank	0.000	00:16
Subject Sample # 2	0.084	00:16
Breath Volume	1.480	00:16
Air Blank	0.000	00:17
Control Test	0.079	00:17
Air Blank	0.000	00:18
Diagnostics Check	OK	00:18

---

Data Download Date/Time: 11/07/2019 15:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARDEN,WILLIAM,L  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: SCHOEN,ELLIS  
ARRESTING OFFICER AGENCY: Winter Garden PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:42  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: RAGER

DATE/TIME OF TEST: 11/02/2019 01:40:52

---

Diagnostics Check	OK	01:34
Air Blank	0.000	01:34
Control Test	0.079	01:35
Air Blank	0.000	01:35
Subject Sample # 1	0.149	01:36
Breath Volume	2.394	01:36
Air Blank	0.000	01:37
Air Blank	0.000	01:38
Subject Sample # 2	0.152	01:39
Breath Volume	2.195	01:39
Air Blank	0.000	01:39
Control Test	0.079	01:40
Air Blank	0.000	01:40
Diagnostics Check	OK	01:40

---

Data Download Date/Time: 11/07/2019 15:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARDEN,WILLIAM,L  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: HARRISON,ROBBIE  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: BAGGS III

DATE/TIME OF TEST: 11/07/2019 00:26:12

---

Diagnostics Check	OK	00:19
Air Blank	0.000	00:20
Control Test	0.079	00:20
Air Blank	0.000	00:21
Subject Sample # 1	0.226	00:21
Breath Volume	2.531	00:21
Air Blank	0.000	00:22
Air Blank	0.000	00:24
Subject Sample # 2	0.229	00:24
Breath Volume	2.113	00:24
Air Blank	0.000	00:25
Control Test	0.078	00:25
Air Blank	0.000	00:26
Diagnostics Check	OK	00:26

---

Data Download Date/Time: 11/07/2019 15:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARDEN,WILLIAM,L  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: HARWOOD,ARMANDO  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: CAPUTO

DATE/TIME OF TEST: 11/08/2019 23:59:52

---

Diagnostics Check	OK	23:53
Air Blank	0.000	23:53
Control Test	0.079	23:54
Air Blank	0.000	23:54
Subject Sample # 1	0.212	23:55
Breath Volume	1.808	23:55
Air Blank	0.000	23:55
Air Blank	0.000	23:57
Subject Sample # 2	0.223	23:58
Breath Volume	2.210	23:58
Air Blank	0.000	23:58
Control Test	0.079	23:59
Air Blank	0.000	23:59
Diagnostics Check	OK	23:59

---

Data Download Date/Time: 11/11/2019 13:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BEEKMAN,JACOB,C  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: KOFFINAS,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: MALDONADO

DATE/TIME OF TEST: 11/01/2019 03:50:38

---

Diagnostics Check	OK	03:42
Air Blank	0.000	03:42
Control Test	0.077	03:42
Air Blank	0.000	03:43
Subject Sample # 1	0.114	03:44
Breath Volume	1.273	03:44
Air Blank	0.000	03:45
Air Blank	0.000	03:47
Subject Sample # 2	0.115	03:49
Breath Volume	1.539	03:49
Air Blank	0.000	03:49
Control Test	0.078	03:50
Air Blank	0.000	03:50
Diagnostics Check	OK	03:50

---

Data Download Date/Time: 11/26/2019 14:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PYE,CHRISTOPHER,T  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: CAMPOS,BARBARA  
ARRESTING OFFICER AGENCY: Kissimmee PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:42  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 11/02/2019 04:56:53

---

Diagnostics Check	OK	04:50
Air Blank	0.000	04:50
Control Test	0.078	04:51
Air Blank	0.000	04:51
Subject Sample # 1	0.166	04:52
Breath Volume	3.332	04:52
Air Blank	0.000	04:53
Air Blank	0.000	04:54
Subject Sample # 2	0.166	04:55
Breath Volume	3.101	04:55
Air Blank	0.000	04:55
Control Test	0.076	04:56
Air Blank	0.000	04:56
Diagnostics Check	OK	04:56

---

Data Download Date/Time: 11/26/2019 14:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PYE,CHRISTOPHER,T  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: PYE,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: RIVERA-QUINONES

DATE/TIME OF TEST: 11/08/2019 03:43:50

---

Diagnostics Check	OK	03:36
Air Blank	0.000	03:36
Control Test	0.078	03:36
Air Blank	0.000	03:37
Subject Sample # 1	0.144	03:38
Breath Volume	1.765	03:38
Air Blank	0.000	03:39
Air Blank	0.000	03:41
Subject Sample # 2	0.143	03:42
Breath Volume	1.773	03:42
Air Blank	0.000	03:42
Control Test	0.078	03:43
Air Blank	0.000	03:43
Diagnostics Check	OK	03:43

---

Data Download Date/Time: 11/26/2019 14:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MORALES, RONNIE,  
BREATH TEST OPERATOR AGENCY: St. Cloud PD  
ARRESTING OFFICER: MORALES, RONNIE  
ARRESTING OFFICER AGENCY: St. Cloud PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 14:04  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 14:08

SUBJECT'S LAST NAME: DURAKIS

DATE/TIME OF TEST: 11/09/2019 15:34:31

---

Diagnostics Check	OK	15:28
Air Blank	0.000	15:29
Control Test	0.080	15:29
Air Blank	0.000	15:29
Subject Sample # 1	0.000	15:30
Breath Volume	2.421	15:30
Air Blank	0.000	15:30
Air Blank	0.000	15:32
Subject Sample # 2	0.000	15:33
Breath Volume	2.742	15:33
Air Blank	0.000	15:33
Control Test	0.079	15:33
Air Blank	0.000	15:34
Diagnostics Check	OK	15:34

---

Data Download Date/Time: 11/26/2019 14:36



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRAWFORD,DAVID,T  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: RODRIGUEZ,ALBERT  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 18:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: GRITZMACHER

DATE/TIME OF TEST: 11/09/2019 19:26:38

---

Diagnostics Check	OK	19:20
Air Blank	0.000	19:20
Control Test	0.077	19:20
Air Blank	0.000	19:21
Subject Sample # 1	0.236	19:22
Breath Volume	2.656	19:22
Air Blank	0.000	19:22
Air Blank	0.000	19:24
Subject Sample # 2	0.247	19:24
Breath Volume	3.109	19:24
Air Blank	0.000	19:25
Control Test	0.075	19:26
Air Blank	0.000	19:26
Diagnostics Check	OK	19:26

---

Data Download Date/Time: 11/26/2019 14:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PYE,CHRISTOPHER,T  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: PYE,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: ROMAN

DATE/TIME OF TEST: 11/11/2019 23:42:05

---

Diagnostics Check	OK	23:36
Air Blank	0.000	23:36
Control Test	0.079	23:36
Air Blank	0.000	23:37
Subject Sample # 1	0.000	23:37
Breath Volume	2.343	23:37
Air Blank	0.000	23:38
Air Blank	0.000	23:40
Subject Sample # 2	0.000	23:40
Breath Volume	2.449	23:40
Air Blank	0.000	23:41
Control Test	0.077	23:41
Air Blank	0.000	23:42
Diagnostics Check	OK	23:42

---

Data Download Date/Time: 11/26/2019 14:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DOUVRES,MICHAEL,J  
BREATH TEST OPERATOR AGENCY: Kissimmee PD  
ARRESTING OFFICER: DOUVRES,MICHAEL  
ARRESTING OFFICER AGENCY: Kissimmee PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: COTTON

DATE/TIME OF TEST: 11/12/2019 00:33:14

---

Diagnostics Check	OK	00:29	
Air Blank	0.000	00:29	
Control Test	0.078	00:30	
Air Blank	0.000	00:30	
Subject Sample	# 1 0.206	00:31	
Breath Volume	1.531	00:31	
Air Blank		00:31	RFI Detect
Air Blank	0.000	00:32	
Control Test	0.078	00:32	
Air Blank	0.000	00:33	
Diagnostics Check	OK	00:33	

---

Data Download Date/Time: 11/26/2019 14:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DOUVRES,MICHAEL,J  
BREATH TEST OPERATOR AGENCY: Kissimmee PD  
ARRESTING OFFICER: DOUVRES,MICHAEL  
ARRESTING OFFICER AGENCY: Kissimmee PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: COTTON

DATE/TIME OF TEST: 11/12/2019 00:43:52

---

Diagnostics Check	OK	00:36
Air Blank	0.000	00:36
Control Test	0.077	00:36
Air Blank	0.000	00:37
Subject Sample # 1	0.216	00:39
Breath Volume	2.890	00:39
Air Blank	0.000	00:40
Air Blank	0.000	00:41
Subject Sample # 2	0.205	00:42
Breath Volume	1.714	00:42
Air Blank	0.000	00:42
Control Test	0.075	00:43
Air Blank	0.000	00:43
Diagnostics Check	OK	00:43

---

Data Download Date/Time: 11/26/2019 14:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AKINS,JUSTIN,M  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: AKINS,JUSTIN  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: SOARES

DATE/TIME OF TEST: 11/12/2019 22:36:40

---

Diagnostics Check	OK	22:30
Air Blank	0.000	22:30
Control Test	0.077	22:31
Air Blank	0.000	22:31
Subject Sample # 1	0.000	22:32
Breath Volume	1.867	22:32
Air Blank	0.000	22:32
Air Blank	0.000	22:34
Subject Sample # 2	0.000	22:35
Breath Volume	1.671	22:35
Air Blank	0.000	22:35
Control Test	0.078	22:36
Air Blank	0.000	22:36
Diagnostics Check	OK	22:36

---

Data Download Date/Time: 11/26/2019 14:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRAWFORD,DAVID,T  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: LATCHMAN,AJYAI  
ARRESTING OFFICER AGENCY: Kissimmee PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 04:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: RIVERA ALICEA

DATE/TIME OF TEST: 11/15/2019 05:16:35

---

Diagnostics Check	OK	05:10	
Air Blank	0.000	05:11	
Control Test	0.080	05:11	
Air Blank	0.000	05:12	
Subject Sample	# 1 0.107	05:15	Volume Not Met
Breath Volume	0.218	05:15	
Air Blank	0.000	05:15	
Air Blank	0.000	05:16	

---

Data Download Date/Time: 11/26/2019 14:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRAWFORD,DAVID,T  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: LATCHMAN,AJYAI  
ARRESTING OFFICER AGENCY: Kissimmee PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 04:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:19

SUBJECT'S LAST NAME: RIVERA ALICEA

DATE/TIME OF TEST: 11/15/2019 05:46:58

---

Diagnostics Check	OK	05:39
Air Blank	0.000	05:40
Control Test	0.079	05:40
Air Blank	0.000	05:40
Subject Sample # 1	0.131	05:42
Breath Volume	1.675	05:42
Air Blank	0.000	05:42
Air Blank	0.000	05:44
Subject Sample # 2	0.132	05:45
Breath Volume	2.097	05:45
Air Blank	0.000	05:46
Control Test	0.078	05:46
Air Blank	0.000	05:46
Diagnostics Check	OK	05:46

---

Data Download Date/Time: 11/26/2019 14:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PANADEIRO,JOSE,L  
BREATH TEST OPERATOR AGENCY: FHP Troop K  
ARRESTING OFFICER: ESCOBAR,RIGOBERTO  
ARRESTING OFFICER AGENCY: FHP Troop K  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 20:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: PANADEIRO

DATE/TIME OF TEST: 11/15/2019 21:40:11

---

Diagnostics Check	OK	21:35	
Air Blank	0.000	21:35	
Control Test	0.080	21:36	
Air Blank	0.000	21:36	
Subject Sample	# 1	21:38	RFI Detect
Breath Volume	0.000	21:38	
Air Blank	0.000	21:39	
Control Test	0.079	21:39	
Air Blank	0.000	21:40	
Diagnostics Check	OK	21:40	

---

Data Download Date/Time: 11/26/2019 14:36



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRAWFORD,DAVID,T  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: CRAWFORD,DAVID  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: GONZALEZ RAMIREZ

DATE/TIME OF TEST: 11/19/2019 01:02:52

---

Diagnostics Check	OK	00:55
Air Blank	0.000	00:56
Control Test	0.080	00:56
Air Blank	0.000	00:57
Subject Sample # 1	0.146	00:57
Breath Volume	1.507	00:57
Air Blank	0.000	00:58
Air Blank	0.000	01:00
Subject Sample # 2	0.145	01:01
Breath Volume	2.332	01:01
Air Blank	0.000	01:01
Control Test	0.079	01:02
Air Blank	0.000	01:02
Diagnostics Check	OK	01:02

---

Data Download Date/Time: 11/26/2019 14:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRAWFORD,DAVID,T  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: CRAWFORD,DAVID  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:01  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: TORRES PIZARRO

DATE/TIME OF TEST: 11/19/2019 21:47:00

---

Diagnostics Check	OK	21:40
Air Blank	0.000	21:41
Control Test	0.080	21:41
Air Blank	0.000	21:41
Subject Sample # 1	0.000	21:42
Breath Volume	2.765	21:42
Air Blank	0.000	21:43
Air Blank	0.000	21:45
Subject Sample # 2	0.000	21:45
Breath Volume	2.402	21:45
Air Blank	0.000	21:46
Control Test	0.079	21:46
Air Blank	0.000	21:46
Diagnostics Check	OK	21:47

---

Data Download Date/Time: 11/26/2019 14:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PYE,CHRISTOPHER,T  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: PYE,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:08  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 11/21/2019 02:51:50

---

Diagnostics Check	OK	02:45
Air Blank	0.000	02:46
Control Test	0.079	02:46
Air Blank	0.000	02:46
Subject Sample # 1	0.000	02:47
Breath Volume	2.046	02:47
Air Blank	0.000	02:47
Air Blank	0.000	02:49
Subject Sample # 2	0.000	02:50
Breath Volume	1.957	02:50
Air Blank	0.000	02:50
Control Test	0.078	02:51
Air Blank	0.000	02:51
Diagnostics Check	OK	02:51

---

Data Download Date/Time: 11/26/2019 14:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FARI,ADAM,N  
BREATH TEST OPERATOR AGENCY: St. Cloud PD  
ARRESTING OFFICER: HELLER,MICHAEL  
ARRESTING OFFICER AGENCY: St. Cloud PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 15:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:13

SUBJECT'S LAST NAME: WOLF

DATE/TIME OF TEST: 11/23/2019 16:55:33

---

Diagnostics Check	OK	16:46
Air Blank	0.000	16:47
Control Test	0.080	16:47
Air Blank	0.000	16:48
Subject Sample # 1	0.071	16:50
Breath Volume	1.476	16:50
Air Blank	0.000	16:51
Air Blank	0.000	16:53
Subject Sample # 2	0.068	16:53
Breath Volume	1.414	16:53
Air Blank	0.000	16:54
Control Test	0.079	16:54
Air Blank	0.000	16:55
Diagnostics Check	OK	16:55

---

Data Download Date/Time: 11/26/2019 14:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HANSELL,ROBERT,E  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: RODRIGUEZ,ALBERT  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: MADERA

DATE/TIME OF TEST: 11/24/2019 04:30:28

---

Diagnostics Check	OK	04:23
Air Blank	0.000	04:24
Control Test	0.079	04:24
Air Blank	0.000	04:25
Subject Sample # 1	0.168	04:25
Breath Volume	5.312	04:25
Air Blank	0.000	04:26
Air Blank	0.000	04:28
Subject Sample # 2	0.157	04:28
Breath Volume	3.617	04:28
Air Blank	0.000	04:29
Control Test	0.077	04:29
Air Blank	0.000	04:30
Diagnostics Check	OK	04:30

---

Data Download Date/Time: 11/26/2019 14:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HANSELL,ROBERT,E  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: SOTO,JOSE  
ARRESTING OFFICER AGENCY: Kissimmee PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 11/24/2019 23:36:15

---

Diagnostics Check	OK	23:30
Air Blank	0.000	23:30
Control Test	0.080	23:30
Air Blank	0.000	23:31
Subject Sample # 1	0.045	23:31
Breath Volume	2.289	23:31
Air Blank	0.000	23:32
Air Blank	0.000	23:34
Subject Sample # 2	0.044	23:34
Breath Volume	2.796	23:34
Air Blank	0.000	23:35
Control Test	0.078	23:35
Air Blank	0.000	23:36
Diagnostics Check	OK	23:36

---

Data Download Date/Time: 11/26/2019 14:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PYE,CHRISTOPHER,T  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: PYE,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 11/25/2019 23:46:53

---

Diagnostics Check	OK	23:41
Air Blank	0.000	23:41
Control Test	0.080	23:41
Air Blank	0.000	23:42
Subject Sample # 1	0.000	23:42
Breath Volume	3.746	23:42
Air Blank	0.000	23:43
Air Blank	0.000	23:45
Subject Sample # 2	0.000	23:45
Breath Volume	3.609	23:45
Air Blank	0.000	23:45
Control Test	0.077	23:46
Air Blank	0.000	23:46
Diagnostics Check	OK	23:46

---

Data Download Date/Time: 11/26/2019 14:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SAYERS,MICHELLE,H  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: SAYERS,MICHELLE  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1190213  
DRY GAS EXPIRY DATE: 08/28/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: CARABALLO COIMBRE

DATE/TIME OF TEST: 11/27/2019 01:08:45

---

Diagnostics Check	OK	01:02
Air Blank	0.000	01:02
Control Test	0.080	01:02
Air Blank	0.000	01:03
Subject Sample # 1	0.140	01:04
Breath Volume	2.117	01:04
Air Blank	0.000	01:04
Air Blank	0.000	01:06
Subject Sample # 2	0.138	01:07
Breath Volume	2.316	01:07
Air Blank	0.000	01:07
Control Test	0.079	01:08
Air Blank	0.000	01:08
Diagnostics Check	OK	01:08

---

Data Download Date/Time: 12/30/2019 16:26



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DOUVRES,MICHAEL,J  
BREATH TEST OPERATOR AGENCY: Kissimmee PD  
ARRESTING OFFICER: CAMPOS,BARBARA  
ARRESTING OFFICER AGENCY: Kissimmee PD  
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1190213  
DRY GAS EXPIRY DATE: 08/28/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: FIGUEROA ANDUJAR

DATE/TIME OF TEST: 11/29/2019 22:44:24

---

Diagnostics Check	OK	22:37
Air Blank	0.000	22:38
Control Test	0.081	22:38
Air Blank	0.000	22:38
Subject Sample # 1	0.048	22:39
Breath Volume	1.472	22:39
Air Blank	0.000	22:40
Air Blank	0.000	22:41
Subject Sample # 2	0.050	22:42
Breath Volume	1.539	22:42
Air Blank	0.000	22:43
Control Test	0.079	22:43
Air Blank	0.000	22:44
Diagnostics Check	OK	22:44

---

Data Download Date/Time: 12/30/2019 16:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PYE,CHRISTOPHER,T  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: PYE,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1190213  
DRY GAS EXPIRY DATE: 08/28/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: RAMIREZ FANTAUZZI

DATE/TIME OF TEST: 11/30/2019 01:02:21

---

Diagnostics Check	OK	00:55
Air Blank	0.000	00:56
Control Test	0.080	00:56
Air Blank	0.000	00:56
Subject Sample # 1	0.138	00:57
Breath Volume	2.375	00:57
Air Blank	0.000	00:58
Air Blank	0.000	01:00
Subject Sample # 2	0.123	01:00
Breath Volume	1.812	01:00
Air Blank	0.000	01:01
Control Test	0.077	01:01
Air Blank	0.000	01:02
Diagnostics Check	OK	01:02

---

Data Download Date/Time: 12/30/2019 16:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SAYERS,MICHELLE,H  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: SAYERS,MICHELLE  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1190213  
DRY GAS EXPIRY DATE: 08/28/2022  
VIOLATION CODE: DUI  
ARREST TIME: 04:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: MADDEN

DATE/TIME OF TEST: 11/30/2019 05:48:18

---

Diagnostics Check	OK	05:41
Air Blank	0.000	05:42
Control Test	0.079	05:42
Air Blank	0.000	05:43
Subject Sample # 1	0.080	05:43
Breath Volume	1.632	05:43
Air Blank	0.000	05:44
Air Blank	0.000	05:46
Subject Sample # 2	0.092	05:46
Breath Volume	1.890	05:46
Air Blank	0.000	05:47
Control Test	0.079	05:47
Air Blank	0.000	05:48
Diagnostics Check	OK	05:48

---

Data Download Date/Time: 12/30/2019 16:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SAYERS,MICHELLE,H  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: SAYERS,MICHELLE  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1190213  
DRY GAS EXPIRY DATE: 08/28/2022  
VIOLATION CODE: DUI  
ARREST TIME: 19:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: ENCARNACION

DATE/TIME OF TEST: 11/30/2019 21:02:17

---

Diagnostics Check	OK	20:56
Air Blank	0.000	20:56
Control Test	0.080	20:56
Air Blank	0.000	20:57
Subject Sample # 1	0.000	20:57
Breath Volume	2.550	20:57
Air Blank	0.000	20:58
Air Blank	0.000	21:00
Subject Sample # 2	0.000	21:00
Breath Volume	1.437	21:00
Air Blank	0.000	21:01
Control Test	0.079	21:01
Air Blank	0.000	21:02
Diagnostics Check	OK	21:02

---

Data Download Date/Time: 12/30/2019 16:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: C650434894500,JOSEPH,N  
BREATH TEST OPERATOR AGENCY: Citrus County SO  
ARRESTING OFFICER: SANTOS,JOSEPH  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: CARINO

DATE/TIME OF TEST: 11/08/2019 02:49:08

---

Diagnostics Check		OK	02:41
Air Blank		0.000	02:42
Control Test		0.078	02:42
Air Blank		0.000	02:43
Subject Sample	# 1	0.085	02:44
Breath Volume		2.792	02:44
Air Blank		0.000	02:44
Air Blank		0.000	02:46
Subject Sample	# 2	0.084	02:47
Breath Volume		2.531	02:47
Air Blank		0.000	02:48
Control Test		0.079	02:48
Air Blank		0.000	02:49
Diagnostics Check		OK	02:49

---

Data Download Date/Time: 11/26/2019 15:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DUNN,COLLIN,J  
BREATH TEST OPERATOR AGENCY: St. Cloud PD  
ARRESTING OFFICER: DUNN,COLLIN  
ARRESTING OFFICER AGENCY: St. Cloud PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:26  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: FRAZIER

DATE/TIME OF TEST: 11/16/2019 00:21:39

---

Diagnostics Check		OK	00:13
Air Blank		0.000	00:13
Control Test		0.079	00:13
Air Blank		0.000	00:14
Subject Sample	# 1	0.222	00:16
Breath Volume		1.234	00:16
Air Blank		0.000	00:16
Air Blank		0.000	00:18
Subject Sample	# 2	0.236	00:20
Breath Volume		2.238	00:20
Air Blank		0.000	00:20
Control Test		0.079	00:21
Air Blank		0.000	00:21
Diagnostics Check		OK	00:21

---

Data Download Date/Time: 11/26/2019 15:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DUNN,COLLIN,J  
BREATH TEST OPERATOR AGENCY: St. Cloud PD  
ARRESTING OFFICER: DUNN,COLLIN  
ARRESTING OFFICER AGENCY: St. Cloud PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: ESCOBAR-CASTANO

DATE/TIME OF TEST: 11/17/2019 04:47:49

---

Diagnostics Check		OK	04:40
Air Blank		0.000	04:41
Control Test		0.079	04:41
Air Blank		0.000	04:42
Subject Sample	# 1	0.102	04:43
Breath Volume		2.273	04:43
Air Blank		0.000	04:43
Air Blank		0.000	04:45
Subject Sample	# 2	0.102	04:46
Breath Volume		2.195	04:46
Air Blank		0.000	04:46
Control Test		0.079	04:47
Air Blank		0.000	04:47
Diagnostics Check		OK	04:47

---

Data Download Date/Time: 11/26/2019 15:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SANTOS,JOSEPH,N  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: SANTOS,JOSEPH  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 11/17/2019 05:25:09

---

Diagnostics Check	OK	05:16
Air Blank	0.000	05:16
Control Test	0.079	05:17
Air Blank	0.000	05:17
Subject Sample # 1	0.131	05:19
Breath Volume	1.269	05:19
Air Blank	0.000	05:20
Air Blank	0.000	05:22
Subject Sample # 2	0.125	05:23
Breath Volume	1.476	05:23
Air Blank	0.000	05:24
Control Test	0.078	05:24
Air Blank	0.000	05:25
Diagnostics Check	OK	05:25

---

Data Download Date/Time: 11/26/2019 15:04



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SAYERS,MICHELLE,H  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: CAMPOS,BARBARA  
ARRESTING OFFICER AGENCY: Kissimmee PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: WILEY

DATE/TIME OF TEST: 11/25/2019 21:13:26

---

Diagnostics Check	OK	21:06
Air Blank	0.000	21:07
Control Test	0.079	21:07
Air Blank	0.000	21:08
Subject Sample # 1	0.266	21:08
Breath Volume	1.457	21:08
Air Blank	0.000	21:09
Air Blank	0.000	21:11
Subject Sample # 2	0.257	21:11
Breath Volume	2.292	21:11
Air Blank	0.000	21:12
Control Test	0.079	21:12
Air Blank	0.000	21:13
Diagnostics Check	OK	21:13

---

Data Download Date/Time: 11/26/2019 15:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EASTMAN,M,R  
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD  
ARRESTING OFFICER: INGERSOLL,N  
ARRESTING OFFICER AGENCY: Cocoa Beach PD  
LAST AGENCY INSPECTION: 10/30/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A3  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: AMULONG

DATE/TIME OF TEST: 11/25/2019 01:26:11

---

Diagnostics Check	OK	01:19
Air Blank	0.000	01:20
Control Test	0.080	01:20
Air Blank	0.000	01:21
Subject Sample # 1	0.097	01:21
Breath Volume	2.828	01:21
Air Blank	0.000	01:22
Air Blank	0.000	01:24
Subject Sample # 2	0.096	01:24
Breath Volume	2.445	01:24
Air Blank	0.000	01:25
Control Test	0.079	01:25
Air Blank	0.000	01:26
Diagnostics Check	OK	01:26

---

Data Download Date/Time: 12/13/2019 14:01

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EASTMAN,M,R  
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD  
ARRESTING OFFICER: PERDUE,J  
ARRESTING OFFICER AGENCY: Cocoa Beach PD  
LAST AGENCY INSPECTION: 10/30/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A3  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:21  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 11/30/2019 03:25:24

---

Diagnostics Check		OK	03:13	
Air Blank		0.000	03:13	
Control Test		0.080	03:14	
Air Blank		0.000	03:14	
Subject Sample	# 1	0.167	03:17	Volume Not Met
Breath Volume		0.304	03:17	
Air Blank		0.000	03:18	
Air Blank		0.000	03:20	
Subject Sample	# 2	0.172	03:20	
Breath Volume		1.636	03:20	
Air Blank		0.000	03:21	
Air Blank		0.000	03:23	
Subject Sample	# 3	0.184	03:23	
Breath Volume		1.996	03:23	
Air Blank		0.000	03:24	
Control Test		0.079	03:24	
Air Blank		0.000	03:25	
Diagnostics Check		OK	03:25	

---

Data Download Date/Time: 12/13/2019 14:01

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SANCHEZ,JASON,M  
BREATH TEST OPERATOR AGENCY: Titusville PD  
ARRESTING OFFICER: QUICK,CHAD  
ARRESTING OFFICER AGENCY: Titusville PD  
LAST AGENCY INSPECTION: 10/29/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29017080A2  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:29  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: BONDARUK

DATE/TIME OF TEST: 11/07/2019 22:33:57

---

Diagnostics Check	OK	22:27
Air Blank	0.000	22:27
Control Test	0.080	22:27
Air Blank	0.000	22:28
Subject Sample # 1	0.214	22:29
Breath Volume	1.226	22:29
Air Blank	0.000	22:29
Air Blank	0.000	22:31
Subject Sample # 2	0.226	22:32
Breath Volume	2.597	22:32
Air Blank	0.000	22:33
Control Test	0.079	22:33
Air Blank	0.000	22:33
Diagnostics Check	OK	22:33

---

Data Download Date/Time: 11/12/2019 09:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DURDEN,SHAWNTRALE,L  
BREATH TEST OPERATOR AGENCY: Titusville PD  
ARRESTING OFFICER: DURDEN,SHAWNTRALE  
ARRESTING OFFICER AGENCY: Titusville PD  
LAST AGENCY INSPECTION: 10/29/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29017080A2  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 16:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:30

SUBJECT'S LAST NAME: LUDDEN

DATE/TIME OF TEST: 11/18/2019 16:22:28

---

Diagnostics Check	OK	16:14
Air Blank	0.000	16:15
Control Test	0.081	16:15
Air Blank	0.000	16:15
Subject Sample # 1	0.150	16:17
Breath Volume	1.539	16:17
Air Blank	0.000	16:18
Air Blank	0.000	16:19
Subject Sample # 2	0.138	16:20
Breath Volume	1.515	16:20
Air Blank	0.000	16:21
Control Test	0.077	16:21
Air Blank	0.000	16:22
Diagnostics Check	OK	16:22

---

Data Download Date/Time: 12/02/2019 04:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HEDERMAN,CASEY,L  
BREATH TEST OPERATOR AGENCY: Titusville PD  
ARRESTING OFFICER: RASMUSSEN,MATTHEW  
ARRESTING OFFICER AGENCY: Titusville PD  
LAST AGENCY INSPECTION: 10/29/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29017080A2  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: PETERS

DATE/TIME OF TEST: 11/24/2019 23:57:59

---

Diagnostics Check	OK	23:51
Air Blank	0.000	23:51
Control Test	0.081	23:51
Air Blank	0.000	23:52
Subject Sample # 1	0.127	23:53
Breath Volume	3.457	23:53
Air Blank	0.000	23:54
Air Blank	0.000	23:55
Subject Sample # 2	0.122	23:56
Breath Volume	2.757	23:56
Air Blank	0.000	23:57
Control Test	0.081	23:57
Air Blank	0.000	23:57
Diagnostics Check	OK	23:57

---

Data Download Date/Time: 12/02/2019 04:40



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HANSON,JASON,A  
BREATH TEST OPERATOR AGENCY: Titusville PD  
ARRESTING OFFICER: ROSA,EMILO  
ARRESTING OFFICER AGENCY: Titusville PD  
LAST AGENCY INSPECTION: 10/29/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29017080A2  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: BUSTELO

DATE/TIME OF TEST: 11/30/2019 02:58:15

---

Diagnostics Check		OK	02:51
Air Blank		0.000	02:52
Control Test		0.080	02:52
Air Blank		0.000	02:53
Subject Sample	# 1	0.162	02:53
Breath Volume		1.226	02:53
Air Blank		0.000	02:54
Air Blank		0.000	02:56
Subject Sample	# 2	0.171	02:56
Breath Volume		1.742	02:56
Air Blank		0.000	02:57
Control Test		0.078	02:57
Air Blank		0.000	02:58
Diagnostics Check		OK	02:58

---

Data Download Date/Time: 12/02/2019 04:40