Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

INSTRUMENT REGISTERED TO: HIALEAH GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUERTA,EDWIN,
BREATH TEST OPERATOR AGENCY: HIALEAH GARDENS PD
ARRESTING OFFICER: HUERTA,EDWIN
ARRESTING OFFICER AGENCY: HIALEAH GARDENS PD
LAST AGENCY INSPECTION: 11/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: MERCADO

DATE/TIME OF TEST: 12/07/2019 02:42:56

Diagnostics Check		OK	02:40	
Air Blank		0.000	02:40	
Control Test		0.078	02:40	
Air Blank		0.000	02:41	
Subject Sample	# 1		02:41	Subject Test Refused
Breath Volume		0.000	02:41	
Air Blank		0.000	02:42	
Control Test		0.078	02:42	
Air Blank		0.000	02:42	
Diagnostics Check		OK	02:42	

Data Download Date/Time: 12/30/2019 10:53

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

INSTRUMENT REGISTERED TO: HIALEAH GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORISMOND,WISNER,
BREATH TEST OPERATOR AGENCY: MEDLEY PD
ARRESTING OFFICER: DORISMOND,WISNER
ARRESTING OFFICER AGENCY: MEDLEY PD
LAST AGENCY INSPECTION: 11/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: LOPEZ LOPEZ

DATE/TIME OF TEST: 12/09/2019 01:05:00

Diagnostics Check		OK	01:02	
Air Blank		0.000	01:02	
Control Test		0.078	01:02	
Air Blank		0.000	01:03	
Subject Sample	# 1		01:03	Subject Test Refused
Breath Volume		0.000	01:03	
Air Blank		0.000	01:04	
Control Test		0.079	01:04	
Air Blank		0.000	01:04	
Diagnostics Check		OK	01:05	

Data Download Date/Time: 12/30/2019 10:53

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

INSTRUMENT REGISTERED TO: HIALEAH GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAVARRIA,DEGLYS,E
BREATH TEST OPERATOR AGENCY: MEDLEY PD
ARRESTING OFFICER: CHAVARRIA,DEGLYS
ARRESTING OFFICER AGENCY: MEDLEY PD
LAST AGENCY INSPECTION: 11/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 12/14/2019 06:04:39

Diagnostics Check		OK	06:01	
Air Blank		0.000	06:02	
Control Test		0.078	06:02	
Air Blank		0.000	06:03	
Subject Sample	# 1		06:03	Subject Test Refused
Breath Volume		0.000	06:03	
Air Blank		0.000	06:03	
Control Test		0.079	06:04	
Air Blank		0.000	06:04	
Diagnostics Check		OK	06:04	

Data Download Date/Time: 12/30/2019 10:53

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

INSTRUMENT REGISTERED TO: HIALEAH GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAVARRIA,DEGLYS,E
BREATH TEST OPERATOR AGENCY: MEDLEY PD
ARRESTING OFFICER: CHAVARRIA,DEGLYS
ARRESTING OFFICER AGENCY: MEDLEY PD
LAST AGENCY INSPECTION: 11/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 06:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:58

SUBJECT'S LAST NAME: ZAMBRANO

DATE/TIME OF TEST: 12/20/2019 07:36:05

Diagnostics Check		OK	07:28
Air Blank		0.000	07:29
Control Test		0.078	07:29
Air Blank		0.000	07:29
Subject Sample	# 1	0.156	07:31
Breath Volume		3.355	07:31
Air Blank		0.000	07:32
Air Blank		0.000	07:34
Subject Sample	# 2	0.152	07:34
Breath Volume		3.472	07:34
Air Blank		0.000	07:35
Control Test		0.078	07:35
Air Blank		0.000	07:35
Diagnostics Check		OK	07:36

Data Download Date/Time: 12/30/2019 10:53

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,D,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: VARGAS,D
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: BERNAL BENGE

DATE/TIME OF TEST: 12/01/2019 03:12:52

Diagnostics Check		OK	03:06
Air Blank		0.000	03:06
Control Test		0.077	03:06
Air Blank		0.000	03:07
Subject Sample	# 1	0.151	03:08
Breath Volume		1.527	03:08
Air Blank		0.000	03:08
Air Blank		0.000	03:10
Subject Sample	# 2	0.139	03:11
Breath Volume		1.773	03:11
Air Blank		0.000	03:11
Control Test		0.077	03:12
Air Blank		0.000	03:12
Diagnostics Check		OK	03:12

Data Download Date/Time: 12/27/2019 11:16

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PACHECO,GEORGE,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: PACHECO,GEORGE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: FORTE

DATE/TIME OF TEST: 12/05/2019 02:22:47

Diagnostics Check		OK	02:19	
Air Blank		0.000	02:20	
Control Test		0.078	02:20	
Air Blank		0.000	02:21	
Subject Sample	# 1		02:21	Subject Test Refused
Breath Volume		0.000	02:21	
Air Blank		0.000	02:21	
Control Test		0.078	02:22	
Air Blank		0.000	02:22	
Diagnostics Check		OK	02:22	

Data Download Date/Time: 12/27/2019 11:16

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: PENA

DATE/TIME OF TEST: 12/06/2019 03:06:43

Diagnostics Check		OK	02:59
Air Blank		0.000	02:59
Control Test		0.078	03:00
Air Blank		0.000	03:00
Subject Sample	# 1	0.165	03:02
Breath Volume		1.746	03:02
Air Blank		0.000	03:02
Air Blank		0.000	03:04
Subject Sample	# 2	0.176	03:05
Breath Volume		1.734	03:05
Air Blank		0.000	03:05
Control Test		0.077	03:06
Air Blank		0.000	03:06
Diagnostics Check		OK	03:06

Data Download Date/Time: 12/27/2019 11:17

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORENZO,YOSNIEL,
BREATH TEST OPERATOR AGENCY: SWEETWATER PD
ARRESTING OFFICER: FERNANDEZ,ORLANDO
ARRESTING OFFICER AGENCY: SWEETWATER PD
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: MOYA HERNANDEZ

DATE/TIME OF TEST: 12/06/2019 03:47:03

Diagnostics Check		OK	03:39
Air Blank		0.000	03:39
Control Test		0.078	03:39
Air Blank		0.000	03:40
Subject Sample	# 1	0.197	03:41
Breath Volume		1.175	03:41
Air Blank		0.000	03:42
Air Blank		0.000	03:43
Subject Sample	# 2	0.192	03:45
Breath Volume		1.363	03:45
Air Blank		0.000	03:46
Control Test		0.079	03:46
Air Blank		0.000	03:46
Diagnostics Check		OK	03:47

Data Download Date/Time: 12/27/2019 11:17

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORENZO,YOSNIEL,
BREATH TEST OPERATOR AGENCY: SWEETWATER PD
ARRESTING OFFICER: LORENZO,YOSNIEL
ARRESTING OFFICER AGENCY: SWEETWATER PD
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 12/07/2019 03:25:54

Diagnostics Check		OK	03:22	
Air Blank		0.000	03:23	
Control Test		0.078	03:23	
Air Blank		0.000	03:24	
Subject Sample	# 1		03:24	Subject Test Refused
Breath Volume		0.000	03:24	
Air Blank		0.000	03:25	
Control Test		0.079	03:25	
Air Blank		0.000	03:25	
Diagnostics Check		OK	03:25	

Data Download Date/Time: 12/27/2019 11:17

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIDDO,RAAJU,S
BREATH TEST OPERATOR AGENCY: OTHER LOCAL
ARRESTING OFFICER: SIDDO,RAAJU
ARRESTING OFFICER AGENCY: OTHER LOCAL
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: BLANDON ZAPATA

DATE/TIME OF TEST: 12/07/2019 04:33:04

Diagnostics Check		OK	04:26
Air Blank		0.000	04:27
Control Test		0.078	04:27
Air Blank		0.000	04:28
Subject Sample	# 1	0.240	04:28
Breath Volume		2.609	04:28
Air Blank		0.000	04:29
Air Blank		0.000	04:31
Subject Sample	# 2	0.252	04:31
Breath Volume		1.847	04:31
Air Blank		0.000	04:32
Control Test		0.077	04:32
Air Blank		0.000	04:32
Diagnostics Check		OK	04:33

Data Download Date/Time: 12/27/2019 11:17

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTANA,JORGE,F
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: QUINTANA,JORGE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: BATAGINI-MAGALHAES

DATE/TIME OF TEST: 12/08/2019 03:58:51

Diagnostics Check		OK	03:52
Air Blank		0.000	03:52
Control Test		0.078	03:53
Air Blank		0.000	03:53
Subject Sample	# 1	0.107	03:54
Breath Volume		3.714	03:54
Air Blank		0.000	03:54
Air Blank		0.000	03:56
Subject Sample	# 2	0.107	03:57
Breath Volume		2.652	03:57
Air Blank		0.000	03:57
Control Test		0.078	03:58
Air Blank		0.000	03:58
Diagnostics Check		OK	03:58

Data Download Date/Time: 12/27/2019 11:17

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PACHECO,GEORGE,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: PACHECO,GEORGE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: QUINTANILLA

DATE/TIME OF TEST: 12/08/2019 05:21:20

Diagnostics Check		OK	05:18	
Air Blank		0.000	05:18	
Control Test		0.078	05:19	
Air Blank		0.000	05:19	
Subject Sample	# 1		05:19	Subject Test Refused
Breath Volume		0.000	05:19	
Air Blank		0.000	05:20	
Control Test		0.078	05:20	
Air Blank		0.000	05:21	
Diagnostics Check		OK	05:21	

Data Download Date/Time: 12/27/2019 11:17

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCAULEY,DEVERO,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: MCAULEY,DEVERO
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: MONTOYA CASTANEDA

DATE/TIME OF TEST: 12/08/2019 23:49:01

Diagnostics Check		OK	23:42
Air Blank		0.000	23:43
Control Test		0.078	23:43
Air Blank		0.000	23:44
Subject Sample	# 1	0.170	23:44
Breath Volume		2.917	23:44
Air Blank		0.000	23:45
Air Blank		0.000	23:47
Subject Sample	# 2	0.165	23:47
Breath Volume		2.167	23:47
Air Blank		0.000	23:48
Control Test		0.078	23:48
Air Blank		0.000	23:48
Diagnostics Check		OK	23:49

Data Download Date/Time: 12/27/2019 11:17

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PACHECO,GEORGE,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: PACHECO,GEORGE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 12/11/2019 02:01:09

Diagnostics Check		OK	01:58	
Air Blank		0.000	01:58	
Control Test		0.078	01:58	
Air Blank		0.000	01:59	
Subject Sample	# 1		01:59	Subject Test Refused
Breath Volume		0.000	01:59	
Air Blank		0.000	02:00	
Control Test		0.078	02:00	
Air Blank		0.000	02:01	
Diagnostics Check		OK	02:01	

Data Download Date/Time: 12/27/2019 11:17

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GATTORNO,ALEXANDER,F
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: GATTORNO,ALEXANDER
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 11:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:27

SUBJECT'S LAST NAME: LIANSK DEGETTO

DATE/TIME OF TEST: 12/11/2019 12:53:36

Diagnostics Check		OK	12:43
Air Blank		0.000	12:44
Control Test		0.078	12:44
Air Blank		0.000	12:45
Subject Sample	# 1	0.162	12:45
Breath Volume		2.605	12:45
Air Blank		0.000	12:46
Air Blank		0.000	12:48
Subject Sample	# 2	0.138	12:48
Breath Volume		1.730	12:48
Air Blank		0.000	12:49
Air Blank		0.000	12:51
Subject Sample	# 3	0.161	12:52
Breath Volume		4.230	12:52
Air Blank		0.000	12:52
Control Test		0.078	12:53
Air Blank		0.000	12:53
Diagnostics Check		OK	12:53

Data Download Date/Time: 12/27/2019 11:17

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,ORLANDO,
BREATH TEST OPERATOR AGENCY: SWEETWATER PD
ARRESTING OFFICER: FERNANDEZ,ORLANDO
ARRESTING OFFICER AGENCY: SWEETWATER PD
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: FERRER

DATE/TIME OF TEST: 12/13/2019 00:44:58

Diagnostics Check		OK	00:42	
Air Blank		0.000	00:42	
Control Test		0.078	00:42	
Air Blank		0.000	00:43	
Subject Sample	# 1		00:43	Subject Test Refused
Breath Volume		0.000	00:43	
Air Blank		0.000	00:44	
Control Test		0.078	00:44	
Air Blank		0.000	00:44	
Diagnostics Check		OK	00:44	

Data Download Date/Time: 12/27/2019 11:17

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTANA,JORGE,F
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: QUINTANA,JORGE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: GUTIERREZ

DATE/TIME OF TEST: 12/14/2019 04:09:34

Diagnostics Check		OK	04:02
Air Blank		0.000	04:03
Control Test		0.078	04:03
Air Blank		0.000	04:04
Subject Sample	# 1	0.125	04:04
Breath Volume		2.300	04:04
Air Blank		0.000	04:05
Air Blank		0.000	04:07
Subject Sample	# 2	0.123	04:08
Breath Volume		1.734	04:08
Air Blank		0.000	04:08
Control Test		0.079	04:08
Air Blank		0.000	04:09
Diagnostics Check		OK	04:09

Data Download Date/Time: 12/27/2019 11:17

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GATTORNO,ALEXANDER,F
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: MURPHY,ANTHONY
ARRESTING OFFICER AGENCY: NORTH MIAMI PD
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: VTOROV

DATE/TIME OF TEST: 12/14/2019 18:45:35

Diagnostics Check		OK	18:35
Air Blank		0.000	18:36
Control Test		0.078	18:36
Air Blank		0.000	18:37
Subject Sample	# 1	0.362	18:37
Breath Volume		3.167	18:37
Air Blank		0.000	18:38
Air Blank		0.000	18:40
Subject Sample	# 2	0.338	18:40
Breath Volume		2.335	18:40
Air Blank		0.000	18:41
Air Blank		0.000	18:43
Subject Sample	# 3	0.357	18:43
Breath Volume		2.992	18:43
Air Blank		0.000	18:44
Control Test		0.077	18:44
Air Blank		0.000	18:45
Diagnostics Check		OK	18:45

Data Download Date/Time: 12/27/2019 11:17

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,ORLANDO,
BREATH TEST OPERATOR AGENCY: SWEETWATER PD
ARRESTING OFFICER: FERNANDEZ,ORLANDO
ARRESTING OFFICER AGENCY: SWEETWATER PD
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: CABRERA RODRIGUEZ

DATE/TIME OF TEST: 12/15/2019 01:58:53

Diagnostics Check		OK	01:52
Air Blank		0.000	01:52
Control Test		0.078	01:53
Air Blank		0.000	01:53
Subject Sample	# 1	0.241	01:54
Breath Volume		1.796	01:54
Air Blank		0.000	01:55
Air Blank		0.000	01:56
Subject Sample	# 2	0.260	01:57
Breath Volume		2.949	01:57
Air Blank		0.000	01:57
Control Test		0.077	01:58
Air Blank		0.000	01:58
Diagnostics Check		OK	01:58

Data Download Date/Time: 12/27/2019 11:17

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCAULEY,DEVERO,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: MCAULEY,DEVERO
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 05:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: RODRIGUEZ PUENTES

DATE/TIME OF TEST: 12/15/2019 05:47:08

Diagnostics Check		OK	05:44	
Air Blank		0.000	05:44	
Control Test		0.078	05:45	
Air Blank		0.000	05:45	
Subject Sample	# 1		05:45	Subject Test Refused
Breath Volume		0.000	05:45	
Air Blank		0.000	05:46	
Control Test		0.078	05:46	
Air Blank		0.000	05:47	
Diagnostics Check		OK	05:47	

Data Download Date/Time: 12/27/2019 11:17

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAINTILIEN,BOB,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: SAINTILIEN,BOB
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: GRAU

DATE/TIME OF TEST: 12/15/2019 22:37:32

Diagnostics Check		OK	22:34	
Air Blank		0.000	22:34	
Control Test		0.078	22:35	
Air Blank		0.000	22:35	
Subject Sample	# 1		22:36	Subject Test Refused
Breath Volume		0.000	22:36	
Air Blank		0.000	22:36	
Control Test		0.078	22:36	
Air Blank		0.000	22:37	
Diagnostics Check		OK	22:37	

Data Download Date/Time: 12/27/2019 11:17

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORENZO,YOSNIEL,
BREATH TEST OPERATOR AGENCY: SWEETWATER PD
ARRESTING OFFICER: AGUILERA,ANDREW
ARRESTING OFFICER AGENCY: OTHER LOCAL
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 12/18/2019 21:52:35

Diagnostics Check		OK	21:48	
Air Blank		0.000	21:49	
Control Test		0.077	21:49	
Air Blank		0.000	21:49	
Subject Sample	# 1		21:51	Subject Test Refused
Breath Volume		0.000	21:51	
Air Blank		0.000	21:51	
Control Test		0.078	21:51	
Air Blank		0.000	21:52	
Diagnostics Check		OK	21:52	

Data Download Date/Time: 12/27/2019 11:17

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,D,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: VARGAS,D
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: MORALES MORALES

DATE/TIME OF TEST: 12/19/2019 04:33:47

Diagnostics Check	OK	04:32	
Air Blank		04:33	Ambient Fail
Air Blank	0.000	04:33	

Data Download Date/Time: 12/27/2019 11:17

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,D,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: VARGAS,D
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: MORALES MORALES

DATE/TIME OF TEST: 12/19/2019 04:42:23

Diagnostics Check		OK	04:36
Air Blank		0.000	04:36
Control Test		0.078	04:36
Air Blank		0.000	04:37
Subject Sample	# 1	0.178	04:37
Breath Volume		3.886	04:37
Air Blank		0.000	04:38
Air Blank		0.000	04:40
Subject Sample	# 2	0.171	04:40
Breath Volume		3.101	04:40
Air Blank		0.000	04:41
Control Test		0.078	04:41
Air Blank		0.000	04:42
Diagnostics Check		OK	04:42

Data Download Date/Time: 12/27/2019 11:17

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUJILLO,GEORGE,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: TRUJILLO,GEORGE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:53

SUBJECT'S LAST NAME: CALIXTE

DATE/TIME OF TEST: 12/19/2019 20:20:39

Diagnostics Check		OK	20:13
Air Blank		0.000	20:14
Control Test		0.078	20:14
Air Blank		0.000	20:14
Subject Sample	# 1	0.088	20:16
Breath Volume		1.332	20:16
Air Blank		0.000	20:17
Air Blank		0.000	20:18
Subject Sample	# 2	0.082	20:19
Breath Volume		1.207	20:19
Air Blank		0.000	20:19
Control Test		0.076	20:20
Air Blank		0.000	20:20
Diagnostics Check		OK	20:20

Data Download Date/Time: 12/27/2019 11:17

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,D,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: VARGAS,D
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: CHATHA

DATE/TIME OF TEST: 12/19/2019 23:50:14

Diagnostics Check		OK	23:43
Air Blank		0.000	23:43
Control Test		0.078	23:44
Air Blank		0.000	23:44
Subject Sample	# 1	0.263	23:45
Breath Volume		1.765	23:45
Air Blank		0.000	23:46
Air Blank		0.000	23:48
Subject Sample	# 2	0.259	23:48
Breath Volume		1.429	23:48
Air Blank		0.000	23:49
Control Test		0.078	23:49
Air Blank		0.000	23:50
Diagnostics Check		OK	23:50

Data Download Date/Time: 12/27/2019 11:17

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PACHECO,GEORGE,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: PACHECO,GEORGE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: VANEGAS

DATE/TIME OF TEST: 12/20/2019 02:39:27

Diagnostics Check		OK	02:36	
Air Blank		0.000	02:36	
Control Test		0.078	02:37	
Air Blank		0.000	02:37	
Subject Sample	# 1		02:38	Subject Test Refused
Breath Volume		0.000	02:38	
Air Blank		0.000	02:38	
Control Test		0.078	02:38	
Air Blank		0.000	02:39	
Diagnostics Check		OK	02:39	

Data Download Date/Time: 12/27/2019 11:17

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PACHECO,GEORGE,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: PACHECO,GEORGE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: FLORES

DATE/TIME OF TEST: 12/21/2019 03:14:29

Diagnostics Check		OK	03:09	
Air Blank		0.000	03:09	
Control Test		0.078	03:09	
Air Blank		0.000	03:10	
Subject Sample	# 1		03:13	Subject Test Refused
Breath Volume		0.000	03:13	
Air Blank		0.000	03:13	
Control Test		0.079	03:13	
Air Blank		0.000	03:14	
Diagnostics Check		OK	03:14	

Data Download Date/Time: 12/27/2019 11:17

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTANA,JORGE,F
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: QUINTANA,JORGE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: ALMARALES

DATE/TIME OF TEST: 12/22/2019 03:12:02

Diagnostics Check		OK	02:59
Air Blank		0.000	02:59
Control Test		0.078	02:59
Air Blank		0.000	03:00
Subject Sample	# 1	0.163	03:03
Breath Volume		1.941	03:03
Air Blank		0.000	03:04
Air Blank		0.000	03:05
Subject Sample	# 2	0.139	03:07
Breath Volume		1.101	03:07
Air Blank		0.000	03:07
Air Blank		0.000	03:09
Subject Sample	# 3	0.167	03:10
Breath Volume		2.113	03:10
Air Blank		0.000	03:11
Control Test		0.077	03:11
Air Blank		0.000	03:11
Diagnostics Check		OK	03:12

Data Download Date/Time: 12/27/2019 11:17

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PACHECO,GEORGE,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: PACHECO,GEORGE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 07:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:53

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 12/25/2019 08:36:52

Diagnostics Check		OK	08:33	
Air Blank		0.000	08:34	
Control Test		0.077	08:34	
Air Blank		0.000	08:35	
Subject Sample	# 1		08:35	Subject Test Refused
Breath Volume		0.000	08:35	
Air Blank		0.000	08:35	
Control Test		0.079	08:36	
Air Blank		0.000	08:36	
Diagnostics Check		OK	08:36	

Data Download Date/Time: 12/27/2019 11:17

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAINTILIEN,BOB,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: SAINTILIEN,BOB
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:21

SUBJECT'S LAST NAME: LOUISSAINT

DATE/TIME OF TEST: 12/26/2019 20:52:56

Diagnostics Check		OK	20:45
Air Blank		0.000	20:46
Control Test		0.077	20:46
Air Blank		0.000	20:47
Subject Sample	# 1	0.167	20:47
Breath Volume		2.570	20:47
Air Blank		0.000	20:48
Air Blank		0.000	20:50
Subject Sample	# 2	0.173	20:51
Breath Volume		1.484	20:51
Air Blank		0.000	20:52
Control Test		0.079	20:52
Air Blank		0.000	20:52
Diagnostics Check		OK	20:52

Data Download Date/Time: 12/27/2019 11:17

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARDELAS,MAYLE,
BREATH TEST OPERATOR AGENCY: OTHER LOCAL
ARRESTING OFFICER: ST THOMAS,JACK
ARRESTING OFFICER AGENCY: OTHER LOCAL
LAST AGENCY INSPECTION: 12/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: CASTAN

DATE/TIME OF TEST: 12/31/2019 01:49:01

Diagnostics Check		OK	01:41
Air Blank		0.000	01:42
Control Test		0.081	01:42
Air Blank		0.000	01:42
Subject Sample	# 1	0.189	01:44
Breath Volume		2.125	01:44
Air Blank		0.000	01:45
Air Blank		0.000	01:46
Subject Sample	# 2	0.193	01:47
Breath Volume		1.156	01:47
Air Blank		0.000	01:48
Control Test		0.077	01:48
Air Blank		0.000	01:48
Diagnostics Check		OK	01:49

Data Download Date/Time: 01/24/2020 09:01

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: GRANDE

DATE/TIME OF TEST: 12/09/2019 03:24:22

Diagnostics Check		OK	03:18
Air Blank		0.000	03:18
Control Test		0.079	03:19
Air Blank		0.000	03:19
Subject Sample	# 1	0.119	03:19
Breath Volume		2.984	03:19
Air Blank		0.000	03:20
Air Blank		0.000	03:22
Subject Sample	# 2	0.116	03:22
Breath Volume		2.152	03:22
Air Blank		0.000	03:23
Control Test		0.078	03:23
Air Blank		0.000	03:24
Diagnostics Check		OK	03:24

Data Download Date/Time: 12/20/2019 10:32

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROOT,K
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: QUACKENBUSH

DATE/TIME OF TEST: 12/11/2019 03:36:43

Diagnostics Check		OK	03:30
Air Blank		0.000	03:30
Control Test		0.079	03:31
Air Blank		0.000	03:31
Subject Sample	# 1	0.178	03:32
Breath Volume		2.949	03:32
Air Blank		0.000	03:32
Air Blank		0.000	03:34
Subject Sample	# 2	0.175	03:35
Breath Volume		1.964	03:35
Air Blank		0.000	03:35
Control Test		0.079	03:36
Air Blank		0.000	03:36
Diagnostics Check		OK	03:36

Data Download Date/Time: 12/20/2019 10:32

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:19

SUBJECT'S LAST NAME: BRYIE

DATE/TIME OF TEST: 12/11/2019 20:46:14

Diagnostics Check		OK	20:40
Air Blank		0.000	20:40
Control Test		0.079	20:40
Air Blank		0.000	20:41
Subject Sample	# 1	0.029	20:41
Breath Volume		2.843	20:41
Air Blank		0.000	20:42
Air Blank		0.000	20:44
Subject Sample	# 2	0.028	20:44
Breath Volume		2.828	20:44
Air Blank		0.000	20:45
Control Test		0.079	20:45
Air Blank		0.000	20:46
Diagnostics Check		OK	20:46

Data Download Date/Time: 12/20/2019 10:32

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: WILLIE

DATE/TIME OF TEST: 12/12/2019 01:38:10

Diagnostics Check		OK	01:31
Air Blank		0.000	01:32
Control Test		0.080	01:32
Air Blank		0.000	01:32
Subject Sample	# 1	0.103	01:33
Breath Volume		1.542	01:33
Air Blank		0.000	01:34
Air Blank		0.000	01:35
Subject Sample	# 2	0.099	01:36
Breath Volume		1.230	01:36
Air Blank		0.000	01:37
Control Test		0.080	01:37
Air Blank		0.000	01:38
Diagnostics Check		OK	01:38

Data Download Date/Time: 12/20/2019 10:32

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,D,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EDWARDS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 07:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:32

SUBJECT'S LAST NAME: BRANDORFF

DATE/TIME OF TEST: 12/16/2019 10:00:18

Diagnostics Check		OK	09:53
Air Blank		0.000	09:53
Control Test		0.080	09:53
Air Blank		0.000	09:54
Subject Sample	# 1	0.218	09:55
Breath Volume		1.812	09:55
Air Blank		0.000	09:56
Air Blank		0.000	09:58
Subject Sample	# 2	0.225	09:58
Breath Volume		3.617	09:58
Air Blank		0.000	09:59
Control Test		0.079	09:59
Air Blank		0.000	10:00
Diagnostics Check		OK	10:00

Data Download Date/Time: 12/20/2019 10:32

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 12/18/2019 02:05:20

Diagnostics Check		OK	01:58
Air Blank		0.000	01:59
Control Test		0.080	01:59
Air Blank		0.000	02:00
Subject Sample	# 1	0.126	02:00
Breath Volume		2.875	02:00
Air Blank		0.000	02:01
Air Blank		0.000	02:03
Subject Sample	# 2	0.121	02:03
Breath Volume		2.671	02:03
Air Blank		0.000	02:04
Control Test		0.079	02:04
Air Blank		0.000	02:05
Diagnostics Check		OK	02:05

Data Download Date/Time: 12/20/2019 10:32

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PAUL,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:25

SUBJECT'S LAST NAME: ANDREWS JR

DATE/TIME OF TEST: 12/18/2019 06:52:21

Diagnostics Check		OK	06:45
Air Blank		0.000	06:46
Control Test		0.080	06:46
Air Blank		0.000	06:47
Subject Sample	# 1	0.062	06:47
Breath Volume		1.156	06:47
Air Blank		0.000	06:48
Air Blank		0.000	06:50
Subject Sample	# 2	0.061	06:50
Breath Volume		1.296	06:50
Air Blank		0.000	06:51
Control Test		0.080	06:51
Air Blank		0.000	06:52
Diagnostics Check		OK	06:52

Data Download Date/Time: 12/20/2019 10:32

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:34

SUBJECT'S LAST NAME: ROSALES CARDENAS

DATE/TIME OF TEST: 12/19/2019 22:10:21

Diagnostics Check		OK	21:55	
Air Blank		0.000	21:56	
Control Test		0.080	21:56	
Air Blank		0.000	21:57	
Subject Sample	# 1	0.193	22:00	Volume Not Met
Breath Volume		1.070	22:00	
Air Blank		0.000	22:00	
Air Blank		0.000	22:02	
Subject Sample	# 2	0.191	22:04	
Breath Volume		1.808	22:04	
Air Blank		0.000	22:05	
Air Blank		0.000	22:07	
Subject Sample	# 3	0.199	22:08	
Breath Volume		1.523	22:08	
Air Blank		0.000	22:09	
Control Test		0.078	22:09	
Air Blank		0.000	22:10	
Diagnostics Check		OK	22:10	

Data Download Date/Time: 12/20/2019 10:32

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TORRES,W
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 12/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: DE ALMEIDA

DATE/TIME OF TEST: 12/20/2019 04:17:57

Diagnostics Check		OK	04:11
Air Blank		0.000	04:11
Control Test		0.079	04:12
Air Blank		0.000	04:12
Subject Sample	# 1	0.166	04:13
Breath Volume		1.726	04:13
Air Blank		0.000	04:13
Air Blank		0.000	04:15
Subject Sample	# 2	0.165	04:16
Breath Volume		3.203	04:16
Air Blank		0.000	04:17
Control Test		0.079	04:17
Air Blank		0.000	04:17
Diagnostics Check		OK	04:17

Data Download Date/Time: 12/20/2019 10:32

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: DOUGLAS

DATE/TIME OF TEST: 12/20/2019 23:00:15

Diagnostics Check		OK	22:52
Air Blank		0.000	22:53
Control Test		0.079	22:53
Air Blank		0.000	22:53
Subject Sample	# 1	0.172	22:55
Breath Volume		1.339	22:55
Air Blank		0.000	22:56
Air Blank		0.000	22:58
Subject Sample	# 2	0.164	22:58
Breath Volume		2.562	22:58
Air Blank		0.000	22:59
Control Test		0.078	22:59
Air Blank		0.000	23:00
Diagnostics Check		OK	23:00

Data Download Date/Time: 01/16/2020 07:45

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: CRABTREE

DATE/TIME OF TEST: 12/21/2019 03:35:52

Diagnostics Check		OK	03:23	
Air Blank		0.000	03:24	
Control Test		0.079	03:24	
Air Blank		0.000	03:24	
Subject Sample	# 1	0.087	03:28	Volume Not Met
Breath Volume		0.613	03:28	
Air Blank		0.000	03:28	
Air Blank		0.000	03:30	
Subject Sample	# 2	0.102	03:31	
Breath Volume		1.910	03:31	
Air Blank		0.000	03:31	
Air Blank		0.000	03:33	
Subject Sample	# 3	0.114	03:34	
Breath Volume		2.425	03:34	
Air Blank		0.000	03:34	
Control Test		0.078	03:35	
Air Blank		0.000	03:35	
Diagnostics Check		OK	03:35	

Data Download Date/Time: 01/16/2020 07:45

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,D,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CAMPBELL,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 11:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:41

SUBJECT'S LAST NAME: DIAZ ARIAS

DATE/TIME OF TEST: 12/21/2019 13:06:37

Diagnostics Check		OK	13:03	
Air Blank		0.000	13:04	
Control Test		0.080	13:04	
Air Blank		0.000	13:05	
Subject Sample	# 1		13:05	Subject Test Refused
Breath Volume		0.000	13:05	
Air Blank		0.000	13:05	
Control Test		0.079	13:06	
Air Blank		0.000	13:06	
Diagnostics Check		OK	13:06	

Data Download Date/Time: 01/16/2020 07:45

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,D,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CRAWFORD,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 15:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:10

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 12/21/2019 17:33:58

Diagnostics Check		OK	17:31	
Air Blank		0.000	17:31	
Control Test		0.080	17:31	
Air Blank		0.000	17:32	
Subject Sample	# 1		17:32	Subject Test Refused
Breath Volume		0.000	17:32	
Air Blank		0.000	17:33	
Control Test		0.079	17:33	
Air Blank		0.000	17:33	
Diagnostics Check		OK	17:33	

Data Download Date/Time: 01/16/2020 07:45

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:23

SUBJECT'S LAST NAME: OWENS

DATE/TIME OF TEST: 12/27/2019 22:49:39

Diagnostics Check		OK	22:43
Air Blank		0.000	22:44
Control Test		0.080	22:44
Air Blank		0.000	22:44
Subject Sample	# 1	0.043	22:45
Breath Volume		1.871	22:45
Air Blank		0.000	22:45
Air Blank		0.000	22:47
Subject Sample	# 2	0.042	22:48
Breath Volume		2.316	22:48
Air Blank		0.000	22:48
Control Test		0.079	22:49
Air Blank		0.000	22:49
Diagnostics Check		OK	22:49

Data Download Date/Time: 01/16/2020 07:45

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TUCKER,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: LYDE

DATE/TIME OF TEST: 12/27/2019 23:34:28

Diagnostics Check		OK	23:28
Air Blank		0.000	23:28
Control Test		0.079	23:28
Air Blank		0.000	23:29
Subject Sample	# 1	0.261	23:30
Breath Volume		1.187	23:30
Air Blank		0.000	23:30
Air Blank		0.000	23:32
Subject Sample	# 2	0.274	23:32
Breath Volume		1.683	23:32
Air Blank		0.000	23:33
Control Test		0.079	23:33
Air Blank		0.000	23:34
Diagnostics Check		OK	23:34

Data Download Date/Time: 01/16/2020 07:45

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: LOVELL

DATE/TIME OF TEST: 12/28/2019 01:28:53

Diagnostics Check		OK	01:22
Air Blank		0.000	01:23
Control Test		0.079	01:23
Air Blank		0.000	01:24
Subject Sample	# 1	0.060	01:24
Breath Volume		3.730	01:24
Air Blank		0.000	01:25
Air Blank		0.000	01:26
Subject Sample	# 2	0.060	01:27
Breath Volume		3.832	01:27
Air Blank		0.000	01:27
Control Test		0.079	01:28
Air Blank		0.000	01:28
Diagnostics Check		OK	01:28

Data Download Date/Time: 01/16/2020 07:45

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: NASH

DATE/TIME OF TEST: 12/28/2019 03:17:07

Diagnostics Check		OK	03:14	
Air Blank		0.000	03:14	
Control Test		0.080	03:14	
Air Blank		0.000	03:15	
Subject Sample	# 1		03:15	Subject Test Refused
Breath Volume		0.000	03:15	
Air Blank		0.000	03:16	
Control Test		0.080	03:16	
Air Blank		0.000	03:17	
Diagnostics Check		OK	03:17	

Data Download Date/Time: 01/16/2020 07:46

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:08

SUBJECT'S LAST NAME: SPORTS

DATE/TIME OF TEST: 12/28/2019 19:35:09

Diagnostics Check		OK	19:28
Air Blank		0.000	19:29
Control Test		0.080	19:29
Air Blank		0.000	19:30
Subject Sample	# 1	0.050	19:30
Breath Volume		2.378	19:30
Air Blank		0.000	19:31
Air Blank		0.000	19:32
Subject Sample	# 2	0.049	19:33
Breath Volume		2.019	19:33
Air Blank		0.000	19:34
Control Test		0.079	19:34
Air Blank		0.000	19:35
Diagnostics Check		OK	19:35

Data Download Date/Time: 01/16/2020 07:46

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:24

SUBJECT'S LAST NAME: BUSHOR

DATE/TIME OF TEST: 12/28/2019 20:50:57

Diagnostics Check		OK	20:44
Air Blank		0.000	20:44
Control Test		0.079	20:44
Air Blank		0.000	20:45
Subject Sample	# 1	0.100	20:46
Breath Volume		1.695	20:46
Air Blank		0.000	20:47
Air Blank		0.000	20:48
Subject Sample	# 2	0.098	20:49
Breath Volume		1.406	20:49
Air Blank		0.000	20:50
Control Test		0.079	20:50
Air Blank		0.000	20:50
Diagnostics Check		OK	20:50

Data Download Date/Time: 01/16/2020 07:46

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: ALEXANDER

DATE/TIME OF TEST: 12/29/2019 01:22:21

Diagnostics Check		OK	01:15
Air Blank		0.000	01:16
Control Test		0.079	01:16
Air Blank		0.000	01:17
Subject Sample	# 1	0.269	01:18
Breath Volume		1.558	01:18
Air Blank		0.000	01:18
Air Blank		0.000	01:20
Subject Sample	# 2	0.261	01:20
Breath Volume		1.199	01:20
Air Blank		0.000	01:21
Control Test		0.078	01:21
Air Blank		0.000	01:22
Diagnostics Check		OK	01:22

Data Download Date/Time: 01/16/2020 07:46

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: WADLINGTON

DATE/TIME OF TEST: 12/29/2019 02:19:39

Diagnostics Check		OK	02:13
Air Blank		0.000	02:13
Control Test		0.080	02:14
Air Blank		0.000	02:14
Subject Sample	# 1	0.094	02:15
Breath Volume		1.164	02:15
Air Blank		0.000	02:15
Air Blank		0.000	02:17
Subject Sample	# 2	0.103	02:18
Breath Volume		1.343	02:18
Air Blank		0.000	02:18
Control Test		0.079	02:19
Air Blank		0.000	02:19
Diagnostics Check		OK	02:19

Data Download Date/Time: 01/16/2020 07:46

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:47

SUBJECT'S LAST NAME: SHELL JR

DATE/TIME OF TEST: 12/29/2019 19:13:46

Diagnostics Check		OK	19:07
Air Blank		0.000	19:08
Control Test		0.079	19:08
Air Blank		0.000	19:08
Subject Sample	# 1	0.115	19:09
Breath Volume		2.851	19:09
Air Blank		0.000	19:09
Air Blank		0.000	19:11
Subject Sample	# 2	0.109	19:12
Breath Volume		2.855	19:12
Air Blank		0.000	19:12
Control Test		0.079	19:13
Air Blank		0.000	19:13
Diagnostics Check		OK	19:13

Data Download Date/Time: 01/16/2020 07:46

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HANKINS,A
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 12/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: DOLPHIN

DATE/TIME OF TEST: 12/30/2019 01:08:13

Diagnostics Check		OK	01:05	
Air Blank		0.000	01:05	
Control Test		0.080	01:06	
Air Blank		0.000	01:06	
Subject Sample	# 1		01:06	Subject Test Refused
Breath Volume		0.000	01:06	
Air Blank		0.000	01:07	
Control Test		0.079	01:07	
Air Blank		0.000	01:08	
Diagnostics Check		OK	01:08	

Data Download Date/Time: 01/16/2020 07:46

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBB,CATHERINE,D
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: BALLARD,J
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: SHIVERS

DATE/TIME OF TEST: 12/01/2019 01:43:28

Diagnostics Check		OK	01:37
Air Blank		0.000	01:37
Control Test		0.079	01:37
Air Blank		0.000	01:38
Subject Sample	# 1	0.032	01:39
Breath Volume		1.507	01:39
Air Blank		0.000	01:39
Air Blank		0.000	01:41
Subject Sample	# 2	0.033	01:42
Breath Volume		1.113	01:42
Air Blank		0.000	01:42
Control Test		0.079	01:42
Air Blank		0.000	01:43
Diagnostics Check		OK	01:43

Data Download Date/Time: 12/31/2019 11:34

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KUTZ,T,E
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: GOENS JR

DATE/TIME OF TEST: 12/03/2019 05:03:36

Diagnostics Check		OK	05:00	
Air Blank		0.000	05:01	
Control Test		0.080	05:01	
Air Blank		0.000	05:02	
Subject Sample	# 1		05:02	Subject Test Refused
Breath Volume		0.000	05:02	
Air Blank		0.000	05:02	
Control Test		0.080	05:02	
Air Blank		0.000	05:03	
Diagnostics Check		OK	05:03	

Data Download Date/Time: 12/31/2019 11:34

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWTON,C,L
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: HARDY

DATE/TIME OF TEST: 12/07/2019 01:24:49

Diagnostics Check		OK	01:18
Air Blank		0.000	01:19
Control Test		0.079	01:19
Air Blank		0.000	01:19
Subject Sample	# 1	0.163	01:20
Breath Volume		1.597	01:20
Air Blank		0.000	01:21
Air Blank		0.000	01:22
Subject Sample	# 2	0.159	01:23
Breath Volume		2.222	01:23
Air Blank		0.000	01:23
Control Test		0.079	01:24
Air Blank		0.000	01:24
Diagnostics Check		OK	01:24

Data Download Date/Time: 12/31/2019 11:34

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWTON,C,L
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: KNIGHT,G
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: FRANCIS

DATE/TIME OF TEST: 12/07/2019 04:02:51

Diagnostics Check		OK	03:54
Air Blank		0.000	03:55
Control Test		0.079	03:55
Air Blank		0.000	03:56
Subject Sample	# 1	0.065	03:58
Breath Volume		1.425	03:58
Air Blank		0.000	03:58
Air Blank		0.000	04:00
Subject Sample	# 2	0.062	04:01
Breath Volume		1.691	04:01
Air Blank		0.000	04:01
Control Test		0.078	04:02
Air Blank		0.000	04:02
Diagnostics Check		OK	04:02

Data Download Date/Time: 12/31/2019 11:34

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWTON,C,L
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: LOVETT

DATE/TIME OF TEST: 12/07/2019 04:57:28

Diagnostics Check		OK	04:49
Air Blank		0.000	04:49
Control Test		0.079	04:49
Air Blank		0.000	04:50
Subject Sample	# 1	0.140	04:52
Breath Volume		1.730	04:52
Air Blank		0.000	04:52
Air Blank		0.000	04:54
Subject Sample	# 2	0.141	04:55
Breath Volume		2.242	04:55
Air Blank		0.000	04:56
Control Test		0.078	04:56
Air Blank		0.000	04:57
Diagnostics Check		OK	04:57

Data Download Date/Time: 12/31/2019 11:34

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KUTZ,T,E
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: BROWNING

DATE/TIME OF TEST: 12/08/2019 02:10:16

Diagnostics Check		OK	02:04
Air Blank		0.000	02:04
Control Test		0.079	02:04
Air Blank		0.000	02:05
Subject Sample	# 1	0.000	02:05
Breath Volume		2.753	02:05
Air Blank		0.000	02:06
Air Blank		0.000	02:08
Subject Sample	# 2	0.000	02:08
Breath Volume		3.347	02:08
Air Blank		0.000	02:09
Control Test		0.077	02:09
Air Blank		0.000	02:10
Diagnostics Check		OK	02:10

Data Download Date/Time: 12/31/2019 11:34

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERON,W,G
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUFFMON,J
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: WOODRUFF

DATE/TIME OF TEST: 12/10/2019 20:48:44

Diagnostics Check		OK	20:41
Air Blank		0.000	20:41
Control Test		0.077	20:42
Air Blank		0.000	20:42
Subject Sample	# 1	0.295	20:43
Breath Volume		2.062	20:43
Air Blank		0.000	20:44
Air Blank		0.000	20:45
Subject Sample	# 2	0.285	20:47
Breath Volume		2.042	20:47
Air Blank		0.000	20:47
Control Test		0.077	20:48
Air Blank		0.000	20:48
Diagnostics Check		OK	20:48

Data Download Date/Time: 12/31/2019 11:34

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMPTON,J,R
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: BAGAN

DATE/TIME OF TEST: 12/12/2019 03:29:51

Diagnostics Check		OK	03:23
Air Blank		0.000	03:24
Control Test		0.079	03:24
Air Blank		0.000	03:24
Subject Sample	# 1	0.000	03:25
Breath Volume		2.488	03:25
Air Blank		0.000	03:26
Air Blank		0.000	03:28
Subject Sample	# 2	0.000	03:28
Breath Volume		2.167	03:28
Air Blank		0.000	03:28
Control Test		0.079	03:29
Air Blank		0.000	03:29
Diagnostics Check		OK	03:29

Data Download Date/Time: 12/31/2019 11:34

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENSA,MICHAEL,
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: YEATMAN,B
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:56

SUBJECT'S LAST NAME: BUCHANAN

DATE/TIME OF TEST: 12/13/2019 19:23:02

Diagnostics Check		OK	19:16
Air Blank		0.000	19:16
Control Test		0.080	19:17
Air Blank		0.000	19:17
Subject Sample	# 1	0.000	19:18
Breath Volume		1.656	19:18
Air Blank		0.000	19:18
Air Blank		0.000	19:20
Subject Sample	# 2	0.000	19:21
Breath Volume		1.660	19:21
Air Blank		0.000	19:22
Control Test		0.078	19:22
Air Blank		0.000	19:22
Diagnostics Check		OK	19:23

Data Download Date/Time: 12/31/2019 11:34

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,MARLON,G
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: BALLARD,J
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 12/14/2019 01:13:10

Diagnostics Check		OK	01:10	
Air Blank		0.000	01:10	
Control Test		0.078	01:11	
Air Blank		0.000	01:11	
Subject Sample	# 1		01:11	Subject Test Refused
Breath Volume		0.000	01:11	
Air Blank		0.000	01:12	
Control Test		0.078	01:12	
Air Blank		0.000	01:13	
Diagnostics Check		OK	01:13	

Data Download Date/Time: 12/31/2019 11:34

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENSA,MICHAEL,
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HALL,J
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: ZUNIGA

DATE/TIME OF TEST: 12/14/2019 02:43:29

Diagnostics Check		OK	02:40	
Air Blank		0.000	02:41	
Control Test		0.078	02:41	
Air Blank		0.000	02:41	
Subject Sample	# 1		02:42	Subject Test Refused
Breath Volume		0.000	02:42	
Air Blank		0.000	02:42	
Control Test		0.079	02:42	
Air Blank		0.000	02:43	
Diagnostics Check		OK	02:43	

Data Download Date/Time: 12/31/2019 11:34

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERON,W,G
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HALL,J
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: DAMICO

DATE/TIME OF TEST: 12/15/2019 01:58:29

Diagnostics Check		OK	01:52
Air Blank		0.000	01:52
Control Test		0.079	01:52
Air Blank		0.000	01:53
Subject Sample	# 1	0.139	01:53
Breath Volume		1.593	01:53
Air Blank		0.000	01:54
Air Blank		0.000	01:56
Subject Sample	# 2	0.127	01:56
Breath Volume		1.273	01:56
Air Blank		0.000	01:57
Control Test		0.078	01:57
Air Blank		0.000	01:58
Diagnostics Check		OK	01:58

Data Download Date/Time: 12/31/2019 11:34

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,DEANNE,R
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUFFMON,J
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 12/15/2019 03:16:08

Diagnostics Check		OK	03:08
Air Blank		0.000	03:09
Control Test		0.079	03:09
Air Blank		0.000	03:09
Subject Sample	# 1	0.098	03:10
Breath Volume		1.242	03:10
Air Blank		0.000	03:11
Air Blank		0.000	03:13
Subject Sample	# 2	0.093	03:14
Breath Volume		1.214	03:14
Air Blank		0.000	03:15
Control Test		0.078	03:15
Air Blank		0.000	03:16
Diagnostics Check		OK	03:16

Data Download Date/Time: 12/31/2019 11:34

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,DEANNE,R
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HALL,J
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: HALLIHAN

DATE/TIME OF TEST: 12/16/2019 00:12:49

Diagnostics Check		OK	00:05
Air Blank		0.000	00:05
Control Test		0.079	00:06
Air Blank		0.000	00:06
Subject Sample	# 1	0.238	00:07
Breath Volume		1.718	00:07
Air Blank		0.000	00:08
Air Blank		0.000	00:10
Subject Sample	# 2	0.224	00:11
Breath Volume		2.164	00:11
Air Blank		0.000	00:11
Control Test		0.078	00:12
Air Blank		0.000	00:12
Diagnostics Check		OK	00:12

Data Download Date/Time: 12/31/2019 11:34

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,J,
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:26

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 12/17/2019 06:56:48

Diagnostics Check		OK	06:50
Air Blank		0.000	06:50
Control Test		0.078	06:51
Air Blank		0.000	06:51
Subject Sample	# 1	0.122	06:52
Breath Volume		1.574	06:52
Air Blank		0.000	06:53
Air Blank		0.000	06:54
Subject Sample	# 2	0.112	06:55
Breath Volume		2.183	06:55
Air Blank		0.000	06:55
Control Test		0.076	06:56
Air Blank		0.000	06:56
Diagnostics Check		OK	06:56

Data Download Date/Time: 12/31/2019 11:34

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENSA,MICHAEL,
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: BALLARD,J
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: BARNES

DATE/TIME OF TEST: 12/19/2019 22:16:16

Diagnostics Check		OK	22:13	
Air Blank		0.000	22:13	
Control Test		0.079	22:14	
Air Blank		0.000	22:14	
Subject Sample	# 1		22:14	Subject Test Refused
Breath Volume		0.000	22:14	
Air Blank		0.000	22:15	
Control Test		0.079	22:15	
Air Blank		0.000	22:16	
Diagnostics Check		OK	22:16	

Data Download Date/Time: 12/31/2019 11:34

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWTON,CARI,L
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 12/21/2019 02:50:48

Diagnostics Check		OK	02:43
Air Blank		0.000	02:44
Control Test		0.078	02:44
Air Blank		0.000	02:45
Subject Sample	# 1	0.106	02:45
Breath Volume		1.632	02:45
Air Blank		0.000	02:46
Air Blank		0.000	02:48
Subject Sample	# 2	0.109	02:49
Breath Volume		1.761	02:49
Air Blank		0.000	02:49
Control Test		0.078	02:50
Air Blank		0.000	02:50
Diagnostics Check		OK	02:50

Data Download Date/Time: 12/31/2019 11:34

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMPTON,JAMAL,R
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: COPHER,CHRISTIAN
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: LINTON

DATE/TIME OF TEST: 12/21/2019 05:22:49

Diagnostics Check		OK	05:16
Air Blank		0.000	05:16
Control Test		0.079	05:17
Air Blank		0.000	05:17
Subject Sample	# 1	0.145	05:18
Breath Volume		3.000	05:18
Air Blank		0.000	05:19
Air Blank		0.000	05:20
Subject Sample	# 2	0.148	05:21
Breath Volume		2.636	05:21
Air Blank		0.000	05:21
Control Test		0.076	05:22
Air Blank		0.000	05:22
Diagnostics Check		OK	05:22

Data Download Date/Time: 12/31/2019 11:34

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCRACKEN,J,D
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME:]

DATE/TIME OF TEST: 12/21/2019 23:00:00

Diagnostics Check		OK	22:57	
Air Blank		0.000	22:57	
Control Test		0.080	22:57	
Air Blank		0.000	22:58	
Subject Sample	# 1		22:58	Subject Test Refused
Breath Volume		0.000	22:58	
Air Blank		0.000	22:59	
Control Test		0.079	22:59	
Air Blank		0.000	22:59	
Diagnostics Check		OK	23:00	

Data Download Date/Time: 12/31/2019 11:34

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMPTON,JAMAL,R
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUNTER,MARK
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 12/22/2019 02:19:10

Diagnostics Check		OK	02:12
Air Blank		0.000	02:13
Control Test		0.079	02:13
Air Blank		0.000	02:13
Subject Sample	# 1	0.117	02:14
Breath Volume		2.281	02:14
Air Blank		0.000	02:15
Air Blank		0.000	02:17
Subject Sample	# 2	0.125	02:17
Breath Volume		3.097	02:17
Air Blank		0.000	02:18
Control Test		0.079	02:18
Air Blank		0.000	02:19
Diagnostics Check		OK	02:19

Data Download Date/Time: 12/31/2019 11:34

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMPTON,JAMAL,R
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: KNIGHT,GEORGE
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 12/22/2019 04:04:43

Diagnostics Check		OK	03:58
Air Blank		0.000	03:58
Control Test		0.079	03:59
Air Blank		0.000	03:59
Subject Sample	# 1	0.177	04:00
Breath Volume		3.078	04:00
Air Blank		0.000	04:00
Air Blank		0.000	04:02
Subject Sample	# 2	0.157	04:03
Breath Volume		3.687	04:03
Air Blank		0.000	04:03
Control Test		0.077	04:04
Air Blank		0.000	04:04
Diagnostics Check		OK	04:04

Data Download Date/Time: 12/31/2019 11:34

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRES,J,L
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: TYSON

DATE/TIME OF TEST: 12/22/2019 22:43:58

Diagnostics Check		OK	22:37
Air Blank		0.000	22:38
Control Test		0.079	22:38
Air Blank		0.000	22:39
Subject Sample	# 1	0.171	22:39
Breath Volume		2.050	22:39
Air Blank		0.000	22:40
Air Blank		0.000	22:42
Subject Sample	# 2	0.166	22:42
Breath Volume		1.863	22:42
Air Blank		0.000	22:43
Control Test		0.079	22:43
Air Blank		0.000	22:43
Diagnostics Check		OK	22:43

Data Download Date/Time: 12/31/2019 11:34

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMPTON,JAMAL,R
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: DUGGAN,L
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: LOWMON

DATE/TIME OF TEST: 12/26/2019 21:29:10

Diagnostics Check		OK	21:23
Air Blank		0.000	21:23
Control Test		0.079	21:23
Air Blank		0.000	21:24
Subject Sample	# 1	0.000	21:24
Breath Volume		4.226	21:24
Air Blank		0.000	21:25
Air Blank		0.000	21:27
Subject Sample	# 2	0.000	21:27
Breath Volume		4.144	21:27
Air Blank		0.000	21:28
Control Test		0.079	21:28
Air Blank		0.000	21:29
Diagnostics Check		OK	21:29

Data Download Date/Time: 12/31/2019 11:34

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,DEANNE,
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: NONE,NONE
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: ADMIN
ARREST TIME: 19:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: WILBURN

DATE/TIME OF TEST: 12/27/2019 19:51:30

Diagnostics Check		OK	19:45
Air Blank		0.000	19:45
Control Test		0.079	19:46
Air Blank		0.000	19:46
Subject Sample	# 1	0.000	19:47
Breath Volume		2.914	19:47
Air Blank		0.000	19:47
Air Blank		0.000	19:49
Subject Sample	# 2	0.000	19:50
Breath Volume		3.105	19:50
Air Blank		0.000	19:50
Control Test		0.079	19:50
Air Blank		0.000	19:51
Diagnostics Check		OK	19:51

Data Download Date/Time: 12/31/2019 11:34

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENSA,MICHAEL,
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: BALLARD,J
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 12:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 12/29/2019 01:56:07

Diagnostics Check		OK	01:49
Air Blank		0.000	01:49
Control Test		0.078	01:50
Air Blank		0.000	01:50
Subject Sample	# 1	0.138	01:51
Breath Volume		2.773	01:51
Air Blank		0.000	01:51
Air Blank		0.000	01:53
Subject Sample	# 2	0.139	01:54
Breath Volume		3.902	01:54
Air Blank		0.000	01:55
Control Test		0.078	01:55
Air Blank		0.000	01:56
Diagnostics Check		OK	01:56

Data Download Date/Time: 12/31/2019 11:34

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMILLAN,SHIKINA,N
BREATH TEST OPERATOR AGENCY: U of Fl PD
ARRESTING OFFICER: WRZESINSKI,TIMOTHY
ARRESTING OFFICER AGENCY: U of Fl PD
LAST AGENCY INSPECTION: 11/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: VARNADORE

DATE/TIME OF TEST: 12/01/2019 05:38:30

Diagnostics Check		OK	05:35	
Air Blank		0.000	05:35	
Control Test		0.078	05:36	
Air Blank		0.000	05:36	
Subject Sample	# 1		05:37	Subject Test Refused
Breath Volume		0.000	05:37	
Air Blank		0.000	05:37	
Control Test		0.077	05:37	
Air Blank		0.000	05:38	
Diagnostics Check		OK	05:38	

Data Download Date/Time: 12/31/2019 02:45

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,RICK,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BROWN,RICK
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 11/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: FONTE

DATE/TIME OF TEST: 12/01/2019 05:50:40

Diagnostics Check		OK	05:43
Air Blank		0.000	05:44
Control Test		0.077	05:44
Air Blank		0.000	05:45
Subject Sample	# 1	0.122	05:46
Breath Volume		1.425	05:46
Air Blank		0.000	05:46
Air Blank		0.000	05:48
Subject Sample	# 2	0.130	05:49
Breath Volume		2.304	05:49
Air Blank		0.000	05:49
Control Test		0.078	05:50
Air Blank		0.000	05:50
Diagnostics Check		OK	05:50

Data Download Date/Time: 12/31/2019 02:45

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001283

INSTRUMENT REGISTERED TO: HIGH SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YAKUBSIN,JAMES,R
BREATH TEST OPERATOR AGENCY: High Springs PD
ARRESTING OFFICER: YAKUBSIN,JAMES
ARRESTING OFFICER AGENCY: High Springs PD
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: CARVER

DATE/TIME OF TEST: 12/01/2019 19:54:53

Diagnostics Check		OK	19:51	
Air Blank		0.000	19:52	
Control Test		0.079	19:52	
Air Blank		0.000	19:53	
Subject Sample	# 1		19:53	Subject Test Refused
Breath Volume		0.000	19:53	
Air Blank		0.000	19:53	
Control Test		0.079	19:54	
Air Blank		0.000	19:54	
Diagnostics Check		OK	19:54	

Data Download Date/Time: 01/08/2020 06:45

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001283

INSTRUMENT REGISTERED TO: HIGH SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMPTON,KENDRICK,L
BREATH TEST OPERATOR AGENCY: High Springs PD
ARRESTING OFFICER: BROWN,TIAUNSHE
ARRESTING OFFICER AGENCY: Alachua PD
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 14:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:52

SUBJECT'S LAST NAME: RYAN

DATE/TIME OF TEST: 12/02/2019 15:09:24

Diagnostics Check		OK	14:57	
Air Blank		0.000	14:58	
Control Test		0.079	14:58	
Air Blank		0.000	14:59	
Subject Sample	# 1		15:02	No Sample Provided
Breath Volume		0.000	15:02	
Air Blank		0.000	15:02	
Air Blank		0.000	15:04	
Subject Sample	# 2		15:08	No Sample Provided
Breath Volume		0.000	15:08	
Air Blank		0.000	15:08	
Control Test		0.080	15:08	
Air Blank		0.000	15:09	
Diagnostics Check		OK	15:09	

Data Download Date/Time: 01/08/2020 06:45

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001283

INSTRUMENT REGISTERED TO: HIGH SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YAKUBSIN,JAMES,R
BREATH TEST OPERATOR AGENCY: High Springs PD
ARRESTING OFFICER: FRICKE,JON
ARRESTING OFFICER AGENCY: High Springs PD
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: BRAUTCHECK

DATE/TIME OF TEST: 12/15/2019 19:18:08

Diagnostics Check		OK	19:11
Air Blank		0.000	19:12
Control Test		0.080	19:12
Air Blank		0.000	19:13
Subject Sample	# 1	0.288	19:13
Breath Volume		3.910	19:13
Air Blank		0.000	19:14
Air Blank		0.000	19:16
Subject Sample	# 2	0.282	19:16
Breath Volume		3.593	19:16
Air Blank		0.000	19:17
Control Test		0.079	19:17
Air Blank		0.000	19:18
Diagnostics Check		OK	19:18

Data Download Date/Time: 01/08/2020 06:45

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

INSTRUMENT REGISTERED TO: BAKER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,C,J
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MILLER,C
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/30/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: TILLMAN

DATE/TIME OF TEST: 12/19/2019 18:12:18

Diagnostics Check		OK	18:05
Air Blank		0.000	18:06
Control Test		0.081	18:06
Air Blank		0.000	18:06
Subject Sample	# 1	0.185	18:07
Breath Volume		1.449	18:07
Air Blank		0.000	18:08
Air Blank		0.000	18:09
Subject Sample	# 2	0.185	18:10
Breath Volume		1.464	18:10
Air Blank		0.000	18:11
Control Test		0.080	18:11
Air Blank		0.000	18:12
Diagnostics Check		OK	18:12

Data Download Date/Time: 12/30/2019 21:18

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

INSTRUMENT REGISTERED TO: BAKER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKLAND,KRIS,M
BREATH TEST OPERATOR AGENCY: Baker County SO
ARRESTING OFFICER: HEATH,MARK
ARRESTING OFFICER AGENCY: Baker County SO
LAST AGENCY INSPECTION: 11/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 12/22/2019 00:56:42

Diagnostics Check		OK	00:50
Air Blank		0.000	00:51
Control Test		0.080	00:51
Air Blank		0.000	00:51
Subject Sample	# 1	0.180	00:52
Breath Volume		3.171	00:52
Air Blank		0.000	00:52
Air Blank		0.000	00:54
Subject Sample	# 2	0.183	00:55
Breath Volume		2.890	00:55
Air Blank		0.000	00:55
Control Test		0.080	00:56
Air Blank		0.000	00:56
Diagnostics Check		OK	00:56

Data Download Date/Time: 12/30/2019 21:18

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001288

INSTRUMENT REGISTERED TO: UNION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOFMANN,TRAVIS,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HOFMANN,TRAVIS
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 823252
DRY GAS EXPIRY DATE: 08/07/2020
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: BIRD

DATE/TIME OF TEST: 12/16/2019 23:13:02

Diagnostics Check		OK	23:10	
Air Blank		0.000	23:10	
Control Test		0.081	23:10	
Air Blank		0.000	23:11	
Subject Sample	# 1		23:11	Subject Test Refused
Breath Volume		0.000	23:11	
Air Blank		0.000	23:12	
Control Test		0.080	23:12	
Air Blank		0.000	23:12	
Diagnostics Check		OK	23:13	

Data Download Date/Time: 01/02/2020 14:04

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,CHARLES,C
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: ROONEY,M
ARRESTING OFFICER AGENCY: Starke PD
LAST AGENCY INSPECTION: 11/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: GINN

DATE/TIME OF TEST: 12/04/2019 23:22:08

Diagnostics Check		OK	23:15
Air Blank		0.000	23:15
Control Test		0.077	23:16
Air Blank		0.000	23:16
Subject Sample	# 1	0.115	23:17
Breath Volume		1.394	23:17
Air Blank		0.000	23:17
Air Blank		0.000	23:19
Subject Sample	# 2	0.120	23:20
Breath Volume		4.093	23:20
Air Blank		0.000	23:21
Control Test		0.078	23:21
Air Blank		0.000	23:22
Diagnostics Check		OK	23:22

Data Download Date/Time: 12/31/2019 09:22

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUMNER,DALTON,R
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: ALSTON,BRANDON
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 11/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 20:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:37

SUBJECT'S LAST NAME: CALI

DATE/TIME OF TEST: 12/20/2019 21:16:49

Diagnostics Check		OK	21:10
Air Blank		0.000	21:10
Control Test		0.078	21:11
Air Blank		0.000	21:11
Subject Sample	# 1	0.252	21:12
Breath Volume		2.078	21:12
Air Blank		0.000	21:12
Air Blank		0.000	21:14
Subject Sample	# 2	0.241	21:15
Breath Volume		1.777	21:15
Air Blank		0.000	21:15
Control Test		0.076	21:16
Air Blank		0.000	21:16
Diagnostics Check		OK	21:16

Data Download Date/Time: 12/31/2019 09:22

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VISMAN,CLAYTON,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: ALSTON,BRANDON
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 11/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 20:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: MASSENGILL

DATE/TIME OF TEST: 12/24/2019 21:02:07

Diagnostics Check		OK	20:55
Air Blank		0.000	20:56
Control Test		0.078	20:56
Air Blank		0.000	20:57
Subject Sample	# 1	0.184	20:57
Breath Volume		2.656	20:57
Air Blank		0.000	20:58
Air Blank		0.000	21:00
Subject Sample	# 2	0.180	21:00
Breath Volume		2.765	21:00
Air Blank		0.000	21:01
Control Test		0.078	21:01
Air Blank		0.000	21:02
Diagnostics Check		OK	21:02

Data Download Date/Time: 12/31/2019 09:22

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUMNER,DALTON,R
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: ALSTON,BRANDON
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 11/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: CASON

DATE/TIME OF TEST: 12/29/2019 01:40:57

Diagnostics Check		OK	01:37	
Air Blank		0.000	01:38	
Control Test		0.079	01:38	
Air Blank		0.000	01:39	
Subject Sample	# 1		01:39	Subject Test Refused
Breath Volume		0.000	01:39	
Air Blank		0.000	01:40	
Control Test		0.079	01:40	
Air Blank		0.000	01:40	
Diagnostics Check		OK	01:40	

Data Download Date/Time: 12/31/2019 09:22

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUMNER,DALTON,R
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: ALSTON,BRANDON
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 11/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: CASON

DATE/TIME OF TEST: 12/29/2019 01:54:08

Diagnostics Check		OK	01:45	
Air Blank		0.000	01:45	
Control Test		0.078	01:45	
Air Blank		0.000	01:46	
Subject Sample	# 1		01:49	No Sample Provided
Breath Volume		0.000	01:49	
Air Blank		0.000	01:49	
Air Blank		0.000	01:51	
Subject Sample	# 2		01:52	Subject Test Refused
Breath Volume		0.000	01:52	
Air Blank		0.000	01:53	
Control Test		0.079	01:53	
Air Blank		0.000	01:54	
Diagnostics Check		OK	01:54	

Data Download Date/Time: 12/31/2019 09:22

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUMNER,DALTON,R
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: ALSTON,BRANDON
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 11/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 12:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: HUTCHINSON

DATE/TIME OF TEST: 12/30/2019 01:45:11

Diagnostics Check		OK	01:37
Air Blank		0.000	01:38
Control Test		0.079	01:38
Air Blank		0.000	01:38
Subject Sample	# 1	0.127	01:40
Breath Volume		1.253	01:40
Air Blank		0.000	01:41
Air Blank		0.000	01:43
Subject Sample	# 2	0.128	01:43
Breath Volume		1.929	01:43
Air Blank		0.000	01:44
Control Test		0.078	01:44
Air Blank		0.000	01:45
Diagnostics Check		OK	01:45

Data Download Date/Time: 12/31/2019 09:22

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,MICHAEL,T
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: THOMAS,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 12/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805701
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 17:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: HERRERA

DATE/TIME OF TEST: 12/22/2019 18:28:46

Diagnostics Check		OK	18:23	
Air Blank		0.000	18:23	
Control Test		0.077	18:23	
Air Blank		0.000	18:24	
Subject Sample	# 1	0.129	18:25	
Breath Volume		2.531	18:25	
Air Blank		0.000	18:25	
Air Blank		0.000	18:27	
Subject Sample	# 2	0.128	18:28	
Breath Volume		2.726	18:28	
Air Blank			18:28	RFI Detect
Air Blank		0.000	18:28	

Data Download Date/Time: 01/18/2020 17:33

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,JAMES,I
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MOORE,JAMES
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1141545
DRY GAS EXPIRY DATE: 05/11/2022
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: ALVARADO ZELAYA

DATE/TIME OF TEST: 12/01/2019 04:19:29

Diagnostics Check		OK	04:16	
Air Blank		0.000	04:16	
Control Test		0.079	04:17	
Air Blank		0.000	04:17	
Subject Sample	# 1		04:18	Subject Test Refused
Breath Volume		0.000	04:18	
Air Blank		0.000	04:18	
Control Test		0.079	04:18	
Air Blank		0.000	04:19	
Diagnostics Check		OK	04:19	

Data Download Date/Time: 12/23/2019 19:09

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER,RAYMOND,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SPENCER,RAYMOND
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1141545
DRY GAS EXPIRY DATE: 05/11/2022
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 12/01/2019 04:43:55

Diagnostics Check		OK	04:40	
Air Blank		0.000	04:40	
Control Test		0.079	04:40	
Air Blank		0.000	04:41	
Subject Sample	# 1		04:42	Subject Test Refused
Breath Volume		0.000	04:42	
Air Blank		0.000	04:43	
Control Test		0.079	04:43	
Air Blank		0.000	04:43	
Diagnostics Check		OK	04:43	

Data Download Date/Time: 12/23/2019 19:09

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER,RAYMOND,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SPENCER,RAYMOND
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1141545
DRY GAS EXPIRY DATE: 05/11/2022
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: STRONG

DATE/TIME OF TEST: 12/01/2019 05:04:53

Diagnostics Check		OK	05:00	
Air Blank		0.000	05:01	
Control Test		0.080	05:01	
Air Blank		0.000	05:02	
Subject Sample	# 1		05:03	Subject Test Refused
Breath Volume		0.000	05:03	
Air Blank		0.000	05:03	
Control Test		0.080	05:04	
Air Blank		0.000	05:04	
Diagnostics Check		OK	05:04	

Data Download Date/Time: 12/23/2019 19:09

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALEANDER,LLOYD,A
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ALEXANDER,LLOYD
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1141545
DRY GAS EXPIRY DATE: 05/11/2022
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: WARD

DATE/TIME OF TEST: 12/04/2019 01:06:02

Diagnostics Check		OK	01:01	
Air Blank		0.000	01:01	
Control Test		0.080	01:01	
Air Blank		0.000	01:02	
Subject Sample	# 1		01:04	Subject Test Refused
Breath Volume		0.289	01:04	
Air Blank		0.000	01:05	
Control Test		0.080	01:05	
Air Blank		0.000	01:05	
Diagnostics Check		OK	01:06	

Data Download Date/Time: 12/23/2019 19:09

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARLSON,JOSHUA,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: AVERA,DANTE
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1141545
DRY GAS EXPIRY DATE: 05/11/2022
VIOLATION CODE: DUI
ARREST TIME: 14:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: PICKETT

DATE/TIME OF TEST: 12/04/2019 15:25:36

Diagnostics Check		OK	15:20	
Air Blank		0.000	15:21	
Control Test		0.080	15:21	
Air Blank		0.000	15:21	
Subject Sample	# 1		15:24	Subject Test Refused
Breath Volume		0.000	15:24	
Air Blank		0.000	15:24	
Control Test		0.080	15:25	
Air Blank		0.000	15:25	
Diagnostics Check		OK	15:25	

Data Download Date/Time: 12/23/2019 19:09

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOYE,JAMES,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MOYE,JAMES
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1141545
DRY GAS EXPIRY DATE: 05/11/2022
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:36

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 12/04/2019 21:04:04

Diagnostics Check		OK	20:57
Air Blank		0.000	20:57
Control Test		0.079	20:58
Air Blank		0.000	20:58
Subject Sample	# 1	0.118	20:59
Breath Volume		3.363	20:59
Air Blank		0.000	21:00
Air Blank		0.000	21:01
Subject Sample	# 2	0.114	21:02
Breath Volume		3.371	21:02
Air Blank		0.000	21:03
Control Test		0.078	21:03
Air Blank		0.000	21:03
Diagnostics Check		OK	21:04

Data Download Date/Time: 12/23/2019 19:09

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOYE,JAMES,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MOYE,JAMES
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1141545
DRY GAS EXPIRY DATE: 05/11/2022
VIOLATION CODE: DUI
ARREST TIME: 20:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:43

SUBJECT'S LAST NAME: KILPATRICK

DATE/TIME OF TEST: 12/05/2019 21:10:07

Diagnostics Check		OK	21:04
Air Blank		0.000	21:04
Control Test		0.080	21:04
Air Blank		0.000	21:05
Subject Sample	# 1	0.000	21:05
Breath Volume		1.894	21:05
Air Blank		0.000	21:06
Air Blank		0.000	21:08
Subject Sample	# 2	0.000	21:08
Breath Volume		1.855	21:08
Air Blank		0.000	21:09
Control Test		0.080	21:09
Air Blank		0.000	21:10
Diagnostics Check		OK	21:10

Data Download Date/Time: 12/23/2019 19:09

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURKE,MATTHEW,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: BURKE,MATTHEW
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1141545
DRY GAS EXPIRY DATE: 05/11/2022
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: WEEKLEY

DATE/TIME OF TEST: 12/06/2019 04:22:04

Diagnostics Check		OK	04:10	
Air Blank		0.000	04:11	
Control Test		0.080	04:11	
Air Blank		0.000	04:11	
Subject Sample	# 1		04:15	No Sample Provided
Breath Volume		0.000	04:15	
Air Blank		0.000	04:15	
Air Blank		0.000	04:17	
Subject Sample	# 2		04:20	No Sample Provided
Breath Volume		0.000	04:20	
Air Blank		0.000	04:21	
Control Test		0.080	04:21	
Air Blank		0.000	04:21	
Diagnostics Check		OK	04:22	

Data Download Date/Time: 12/23/2019 19:09

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORENE,AIDAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: NORENE,AIDAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1141545
DRY GAS EXPIRY DATE: 05/11/2022
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: CAMACHO JUSINO

DATE/TIME OF TEST: 12/07/2019 01:27:09

Diagnostics Check		OK	01:20
Air Blank		0.000	01:20
Control Test		0.080	01:21
Air Blank		0.000	01:21
Subject Sample	# 1	0.143	01:22
Breath Volume		2.207	01:22
Air Blank		0.000	01:22
Air Blank		0.000	01:24
Subject Sample	# 2	0.144	01:25
Breath Volume		2.550	01:25
Air Blank		0.000	01:26
Control Test		0.080	01:26
Air Blank		0.000	01:27
Diagnostics Check		OK	01:27

Data Download Date/Time: 12/23/2019 19:09

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUNTER,REGEN,B
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: GUNTER,REGEN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1141545
DRY GAS EXPIRY DATE: 05/11/2022
VIOLATION CODE: DUI
ARREST TIME: 15:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:01

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 12/07/2019 15:49:24

Diagnostics Check		OK	15:42
Air Blank		0.000	15:43
Control Test		0.080	15:43
Air Blank		0.000	15:43
Subject Sample	# 1	0.102	15:44
Breath Volume		1.796	15:44
Air Blank		0.000	15:45
Air Blank		0.000	15:47
Subject Sample	# 2	0.099	15:47
Breath Volume		3.558	15:47
Air Blank		0.000	15:48
Control Test		0.079	15:48
Air Blank		0.000	15:49
Diagnostics Check		OK	15:49

Data Download Date/Time: 12/23/2019 19:09

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUNTER,REGEN,B
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: GUNTER,REGEN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1141545
DRY GAS EXPIRY DATE: 05/11/2022
VIOLATION CODE: DUI
ARREST TIME: 19:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: REBMAN

DATE/TIME OF TEST: 12/07/2019 20:27:26

Diagnostics Check		OK	20:20
Air Blank		0.000	20:20
Control Test		0.080	20:21
Air Blank		0.000	20:21
Subject Sample	# 1	0.152	20:22
Breath Volume		2.886	20:22
Air Blank		0.000	20:23
Air Blank		0.000	20:25
Subject Sample	# 2	0.154	20:25
Breath Volume		3.289	20:25
Air Blank		0.000	20:26
Control Test		0.079	20:26
Air Blank		0.000	20:27
Diagnostics Check		OK	20:27

Data Download Date/Time: 12/23/2019 19:09

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAPMAN,MAVERICK,E
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CHAPMAN,MAVERICK
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1141545
DRY GAS EXPIRY DATE: 05/11/2022
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: PERRITT

DATE/TIME OF TEST: 12/08/2019 01:45:12

Diagnostics Check		OK	01:42	
Air Blank		0.000	01:42	
Control Test		0.080	01:42	
Air Blank		0.000	01:43	
Subject Sample	# 1		01:43	Subject Test Refused
Breath Volume		0.000	01:43	
Air Blank		0.000	01:44	
Control Test		0.080	01:44	
Air Blank		0.000	01:45	
Diagnostics Check		OK	01:45	

Data Download Date/Time: 12/23/2019 19:09

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKSON,DAKOTA,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: LYNN,JAMISON
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1141545
DRY GAS EXPIRY DATE: 05/11/2022
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: SALAZAR

DATE/TIME OF TEST: 12/09/2019 23:19:10

Diagnostics Check		OK	23:10	
Air Blank		0.000	23:10	
Control Test		0.080	23:11	
Air Blank		0.000	23:11	
Subject Sample	# 1		23:14	No Sample Provided
Breath Volume		0.000	23:14	
Air Blank		0.000	23:15	
Air Blank		0.000	23:17	
Subject Sample	# 2		23:17	Subject Test Refused
Breath Volume		0.000	23:17	
Air Blank		0.000	23:18	
Control Test		0.080	23:18	
Air Blank		0.000	23:19	
Diagnostics Check		OK	23:19	

Data Download Date/Time: 12/23/2019 19:09

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREIRA,JOSE,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: PEREIRA,JOSE
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1141545
DRY GAS EXPIRY DATE: 05/11/2022
VIOLATION CODE: DUI
ARREST TIME: 13:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:41

SUBJECT'S LAST NAME: COTTEN

DATE/TIME OF TEST: 12/10/2019 14:12:03

Diagnostics Check		OK	14:02	
Air Blank		0.000	14:02	
Control Test		0.079	14:03	
Air Blank		0.000	14:03	
Subject Sample	# 1		14:06	No Sample Provided
Breath Volume		0.000	14:06	
Air Blank		0.000	14:07	
Air Blank		0.000	14:09	
Subject Sample	# 2		14:10	Subject Test Refused
Breath Volume		0.000	14:10	
Air Blank		0.000	14:11	
Control Test		0.080	14:11	
Air Blank		0.000	14:11	
Diagnostics Check		OK	14:12	

Data Download Date/Time: 12/23/2019 19:09

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LARRIVEE-GRANT,CHRISTINA,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: LARRIVEE-GRANT,CHRISTINA
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1141545
DRY GAS EXPIRY DATE: 05/11/2022
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: RUSSELL

DATE/TIME OF TEST: 12/12/2019 02:19:43

Diagnostics Check		OK	02:14	
Air Blank		0.000	02:15	
Control Test		0.080	02:15	
Air Blank		0.000	02:15	
Subject Sample	# 1	0.000	02:16	
Breath Volume		2.050	02:16	
Air Blank		0.000	02:16	
Air Blank			02:18	Improper Sample
Air Blank		0.000	02:18	
Control Test		0.081	02:19	
Air Blank		0.000	02:19	
Diagnostics Check		OK	02:19	

Data Download Date/Time: 12/23/2019 19:09

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LARIVEE-GRANT,CHRISTINA,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: LARRIVEE-GRANT,CHRISTINA
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1141545
DRY GAS EXPIRY DATE: 05/11/2022
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: RUSSELL

DATE/TIME OF TEST: 12/12/2019 02:29:23

Diagnostics Check		OK	02:23
Air Blank		0.000	02:23
Control Test		0.081	02:24
Air Blank		0.000	02:24
Subject Sample	# 1	0.000	02:25
Breath Volume		1.910	02:25
Air Blank		0.000	02:25
Air Blank		0.000	02:27
Subject Sample	# 2	0.000	02:28
Breath Volume		1.863	02:28
Air Blank		0.000	02:28
Control Test		0.079	02:28
Air Blank		0.000	02:29
Diagnostics Check		OK	02:29

Data Download Date/Time: 12/23/2019 19:10

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAPMAN,MAVERICK,E
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: LYNN,JAMISON
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1141545
DRY GAS EXPIRY DATE: 05/11/2022
VIOLATION CODE: DUI
ARREST TIME: 20:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:08

SUBJECT'S LAST NAME: STOTT

DATE/TIME OF TEST: 12/14/2019 21:31:40

Diagnostics Check		OK	21:28	
Air Blank		0.000	21:29	
Control Test		0.080	21:29	
Air Blank		0.000	21:29	
Subject Sample	# 1		21:30	Subject Test Refused
Breath Volume		0.000	21:30	
Air Blank		0.000	21:30	
Control Test		0.080	21:31	
Air Blank		0.000	21:31	
Diagnostics Check		OK	21:31	

Data Download Date/Time: 12/23/2019 19:10

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: APPLING,JEFF,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: APPLING,JEFF
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1141545
DRY GAS EXPIRY DATE: 05/11/2022
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: DICKERSON

DATE/TIME OF TEST: 12/15/2019 01:18:39

Diagnostics Check		OK	01:12
Air Blank		0.000	01:12
Control Test		0.080	01:13
Air Blank		0.000	01:13
Subject Sample	# 1	0.169	01:14
Breath Volume		1.859	01:14
Air Blank		0.000	01:14
Air Blank		0.000	01:16
Subject Sample	# 2	0.159	01:17
Breath Volume		1.480	01:17
Air Blank		0.000	01:17
Control Test		0.081	01:18
Air Blank		0.000	01:18
Diagnostics Check		OK	01:18

Data Download Date/Time: 12/23/2019 19:10

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNSFORD,WILLIAM,C
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: DUNSFORD,WILLIAM
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1141545
DRY GAS EXPIRY DATE: 05/11/2022
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: NGUYEN

DATE/TIME OF TEST: 12/15/2019 02:28:44

Diagnostics Check		OK	02:20	
Air Blank		0.000	02:20	
Control Test		0.081	02:20	
Air Blank		0.000	02:21	
Subject Sample	# 1		02:24	No Sample Provided
Breath Volume		0.000	02:24	
Air Blank		0.000	02:25	
Air Blank		0.000	02:27	
Subject Sample	# 2		02:27	Subject Test Refused
Breath Volume		0.000	02:27	
Air Blank		0.000	02:27	
Control Test		0.080	02:28	
Air Blank		0.000	02:28	
Diagnostics Check		OK	02:28	

Data Download Date/Time: 12/23/2019 19:10

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER,RAYMOND,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SPENCER,RAYMOND
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1141545
DRY GAS EXPIRY DATE: 05/11/2022
VIOLATION CODE: DUI
ARREST TIME: 04:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:20

SUBJECT'S LAST NAME: JEFFRIES

DATE/TIME OF TEST: 12/15/2019 06:48:37

Diagnostics Check		OK	06:41
Air Blank		0.000	06:42
Control Test		0.080	06:42
Air Blank		0.000	06:42
Subject Sample	# 1	0.056	06:43
Breath Volume		2.601	06:43
Air Blank		0.000	06:44
Air Blank		0.000	06:46
Subject Sample	# 2	0.053	06:47
Breath Volume		1.515	06:47
Air Blank		0.000	06:47
Control Test		0.080	06:48
Air Blank		0.000	06:48
Diagnostics Check		OK	06:48

Data Download Date/Time: 12/23/2019 19:10

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEICHNER,ADAM,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: TEICHNER,ADAM
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1141545
DRY GAS EXPIRY DATE: 05/11/2022
VIOLATION CODE: DUI
ARREST TIME: 13:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:13

SUBJECT'S LAST NAME: MATZINGER

DATE/TIME OF TEST: 12/15/2019 14:43:19

Diagnostics Check		OK	14:33
Air Blank		0.000	14:34
Control Test		0.079	14:34
Air Blank		0.000	14:35
Subject Sample	# 1	0.301	14:35
Breath Volume		3.281	14:35
Air Blank		0.000	14:36
Air Blank		0.000	14:38
Subject Sample	# 2	0.279	14:38
Breath Volume		1.894	14:38
Air Blank		0.000	14:39
Air Blank		0.000	14:41
Subject Sample	# 3	0.293	14:41
Breath Volume		3.546	14:41
Air Blank		0.000	14:42
Control Test		0.078	14:42
Air Blank		0.000	14:43
Diagnostics Check		OK	14:43

Data Download Date/Time: 12/23/2019 19:10

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,KAYLAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: DAVIS,KAYLAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1141545
DRY GAS EXPIRY DATE: 05/11/2022
VIOLATION CODE: DUI
ARREST TIME: 16:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:19

SUBJECT'S LAST NAME: ERWIN

DATE/TIME OF TEST: 12/16/2019 18:45:32

Diagnostics Check		OK	18:39
Air Blank		0.000	18:39
Control Test		0.079	18:40
Air Blank		0.000	18:40
Subject Sample	# 1	0.233	18:41
Breath Volume		1.261	18:41
Air Blank		0.000	18:41
Air Blank		0.000	18:43
Subject Sample	# 2	0.222	18:44
Breath Volume		2.285	18:44
Air Blank		0.000	18:44
Control Test		0.079	18:44
Air Blank		0.000	18:45
Diagnostics Check		OK	18:45

Data Download Date/Time: 12/23/2019 19:10

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: APPLING,JEFF,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: APPILNG,JEFF
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1141545
DRY GAS EXPIRY DATE: 05/11/2022
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: DURDEN

DATE/TIME OF TEST: 12/20/2019 21:48:54

Diagnostics Check		OK	21:42
Air Blank		0.000	21:43
Control Test		0.080	21:43
Air Blank		0.000	21:44
Subject Sample	# 1	0.155	21:44
Breath Volume		2.464	21:44
Air Blank		0.000	21:45
Air Blank		0.000	21:46
Subject Sample	# 2	0.160	21:47
Breath Volume		2.261	21:47
Air Blank		0.000	21:48
Control Test		0.080	21:48
Air Blank		0.000	21:48
Diagnostics Check		OK	21:48

Data Download Date/Time: 12/23/2019 19:10

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVIER,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: LEVIER,MATTHEW
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1141545
DRY GAS EXPIRY DATE: 05/11/2022
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: LOCKLIN

DATE/TIME OF TEST: 12/21/2019 02:17:44

Diagnostics Check		OK	02:11
Air Blank		0.000	02:11
Control Test		0.081	02:12
Air Blank		0.000	02:12
Subject Sample	# 1	0.117	02:13
Breath Volume		2.050	02:13
Air Blank		0.000	02:13
Air Blank		0.000	02:15
Subject Sample	# 2	0.111	02:16
Breath Volume		1.976	02:16
Air Blank		0.000	02:16
Control Test		0.080	02:17
Air Blank		0.000	02:17
Diagnostics Check		OK	02:17

Data Download Date/Time: 12/23/2019 19:10

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER,RAYMOND,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SPENCER,RAYMOND
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1141545
DRY GAS EXPIRY DATE: 05/11/2022
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: LEAVITT

DATE/TIME OF TEST: 12/21/2019 04:14:43

Diagnostics Check		OK	04:05
Air Blank		0.000	04:05
Control Test		0.080	04:05
Air Blank		0.000	04:06
Subject Sample	# 1	0.153	04:07
Breath Volume		1.464	04:07
Air Blank		0.000	04:08
Air Blank		0.000	04:09
Subject Sample	# 2	0.160	04:13
Breath Volume		1.976	04:13
Air Blank		0.000	04:13
Control Test		0.081	04:14
Air Blank		0.000	04:14
Diagnostics Check		OK	04:14

Data Download Date/Time: 12/23/2019 19:10

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001307

INSTRUMENT REGISTERED TO: MADISON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VOLLERTSEN,CRAIG,W
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: VOLLERTSEN,CRAIG
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 12/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: TYLER

DATE/TIME OF TEST: 12/14/2019 02:51:31

Diagnostics Check		OK	02:45
Air Blank		0.000	02:45
Control Test		0.080	02:46
Air Blank		0.000	02:46
Subject Sample	# 1	0.067	02:47
Breath Volume		2.828	02:47
Air Blank		0.000	02:47
Air Blank		0.000	02:49
Subject Sample	# 2	0.058	02:50
Breath Volume		1.703	02:50
Air Blank		0.000	02:50
Control Test		0.079	02:50
Air Blank		0.000	02:51
Diagnostics Check		OK	02:51

Data Download Date/Time: 01/06/2020 13:58

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001307

INSTRUMENT REGISTERED TO: MADISON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELLS,DEONSHAY,D
BREATH TEST OPERATOR AGENCY: Madison County SO
ARRESTING OFFICER: SANDERS,ROBERT
ARRESTING OFFICER AGENCY: Madison PD
LAST AGENCY INSPECTION: 12/09/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: WELLS

DATE/TIME OF TEST: 12/22/2019 05:32:18

Diagnostics Check		OK	05:26
Air Blank		0.000	05:26
Control Test		0.081	05:27
Air Blank		0.000	05:27
Subject Sample	# 1	0.141	05:27
Breath Volume		1.722	05:27
Air Blank		0.000	05:28
Air Blank		0.000	05:30
Subject Sample	# 2	0.148	05:30
Breath Volume		2.339	05:30
Air Blank		0.000	05:31
Control Test		0.079	05:31
Air Blank		0.000	05:32
Diagnostics Check		OK	05:32

Data Download Date/Time: 01/06/2020 13:58

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001307

INSTRUMENT REGISTERED TO: MADISON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARREN,BRIAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: WARREN,BRIAN
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 12/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 08:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:27

SUBJECT'S LAST NAME: QUINN

DATE/TIME OF TEST: 12/25/2019 09:51:37

Diagnostics Check		OK	09:48	
Air Blank		0.000	09:49	
Control Test		0.081	09:49	
Air Blank		0.000	09:49	
Subject Sample	# 1		09:50	RFI Detect
Breath Volume		0.000	09:50	
Air Blank		0.000	09:50	
Control Test		0.082	09:51	
Air Blank		0.000	09:51	
Diagnostics Check		OK	09:51	

Data Download Date/Time: 01/06/2020 13:58

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001307

INSTRUMENT REGISTERED TO: MADISON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARREN,BRIAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: WARREN,BRIAN
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 12/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 08:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:27

SUBJECT'S LAST NAME: QUINN

DATE/TIME OF TEST: 12/25/2019 09:56:29

Diagnostics Check		OK	09:53	
Air Blank		0.000	09:54	
Control Test		0.081	09:54	
Air Blank		0.000	09:54	
Subject Sample	# 1		09:55	RFI Detect
Breath Volume		0.000	09:55	
Air Blank		0.000	09:55	
Control Test		0.081	09:55	
Air Blank		0.000	09:56	
Diagnostics Check		OK	09:56	

Data Download Date/Time: 01/06/2020 13:58

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAWYER,DARRYL,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: DELANY,TYLER
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: HARDESTY

DATE/TIME OF TEST: 12/01/2019 02:03:58

Diagnostics Check		OK	01:57
Air Blank		0.000	01:58
Control Test		0.081	01:58
Air Blank		0.000	01:58
Subject Sample	# 1	0.030	01:59
Breath Volume		1.199	01:59
Air Blank		0.000	02:00
Air Blank		0.000	02:02
Subject Sample	# 2	0.030	02:02
Breath Volume		1.589	02:02
Air Blank		0.000	02:03
Control Test		0.081	02:03
Air Blank		0.000	02:03
Diagnostics Check		OK	02:03

Data Download Date/Time: 12/23/2019 11:39

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MERRITT,DAKOTA,A
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: ANDERSON,JOSHUA
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: TINGSTROM

DATE/TIME OF TEST: 12/07/2019 21:41:27

Diagnostics Check		OK	21:33
Air Blank		0.000	21:34
Control Test		0.081	21:34
Air Blank		0.000	21:34
Subject Sample	# 1	0.000	21:36
Breath Volume		1.511	21:36
Air Blank		0.000	21:36
Air Blank		0.000	21:38
Subject Sample	# 2	0.000	21:40
Breath Volume		1.761	21:40
Air Blank		0.000	21:40
Control Test		0.081	21:40
Air Blank		0.000	21:41
Diagnostics Check		OK	21:41

Data Download Date/Time: 12/23/2019 11:39

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WITHROW,TRACY,L
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: WITHROW,TRACY
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: ANDREWS

DATE/TIME OF TEST: 12/12/2019 05:22:47

Diagnostics Check		OK	05:11	
Air Blank		0.000	05:11	
Control Test		0.080	05:11	
Air Blank		0.000	05:12	
Subject Sample	# 1	0.121	05:15	Volume Not Met
Breath Volume		0.187	05:15	
Air Blank		0.000	05:16	
Air Blank		0.000	05:18	
Subject Sample	# 2	0.136	05:21	Volume Not Met
Breath Volume		0.394	05:21	
Air Blank		0.000	05:21	
Control Test		0.081	05:22	
Air Blank		0.000	05:22	
Diagnostics Check		OK	05:22	

Data Download Date/Time: 12/23/2019 11:40

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WITHROW,TRACY,L
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: WITHROW,TRACY
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: ANDREWS

DATE/TIME OF TEST: 12/12/2019 05:29:38

Diagnostics Check		OK	05:25	
Air Blank		0.000	05:25	
Control Test		0.079	05:25	
Air Blank		0.000	05:26	
Subject Sample	# 1		05:28	Subject Test Refused
Breath Volume		1.187	05:28	
Air Blank		0.000	05:28	
Control Test		0.080	05:29	
Air Blank		0.000	05:29	
Diagnostics Check		OK	05:29	

Data Download Date/Time: 12/23/2019 11:40

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,TIMOTHY,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LEWIS,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 14:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:26

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 12/15/2019 15:53:53

Diagnostics Check		OK	15:47
Air Blank		0.000	15:47
Control Test		0.080	15:48
Air Blank		0.000	15:48
Subject Sample	# 1	0.305	15:49
Breath Volume		1.894	15:49
Air Blank		0.000	15:49
Air Blank		0.000	15:51
Subject Sample	# 2	0.307	15:52
Breath Volume		1.148	15:52
Air Blank		0.000	15:52
Control Test		0.080	15:53
Air Blank		0.000	15:53
Diagnostics Check		OK	15:53

Data Download Date/Time: 12/23/2019 11:40

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIDYETTE,COREY,L
BREATH TEST OPERATOR AGENCY: Lynn Haven PD
ARRESTING OFFICER: CONLEY,ALISHA
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 18:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: CARPENTER

DATE/TIME OF TEST: 12/19/2019 19:03:51

Diagnostics Check		OK	19:01	
Air Blank		0.000	19:01	
Control Test		0.081	19:01	
Air Blank		0.000	19:02	
Subject Sample	# 1		19:02	Subject Test Refused
Breath Volume		0.000	19:02	
Air Blank		0.000	19:02	
Control Test		0.080	19:03	
Air Blank		0.000	19:03	
Diagnostics Check		OK	19:03	

Data Download Date/Time: 12/23/2019 11:40

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER,MONTREZ,R
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: POTTER,MONTREZ
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 12/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 19:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:47

SUBJECT'S LAST NAME: HOLLAND

DATE/TIME OF TEST: 12/23/2019 20:10:00

Diagnostics Check		OK	20:07	
Air Blank		0.000	20:07	
Control Test		0.081	20:07	
Air Blank		0.000	20:08	
Subject Sample	# 1		20:08	Subject Test Refused
Breath Volume		0.000	20:08	
Air Blank		0.000	20:09	
Control Test		0.081	20:09	
Air Blank		0.000	20:09	
Diagnostics Check		OK	20:10	

Data Download Date/Time: 01/30/2020 14:57

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,TIMOTHY,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LEWIS,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 20:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: GUNAWAN

DATE/TIME OF TEST: 12/28/2019 22:03:21

Diagnostics Check		OK	21:57
Air Blank		0.000	21:57
Control Test		0.081	21:57
Air Blank		0.000	21:58
Subject Sample	# 1	0.000	21:59
Breath Volume		1.742	21:59
Air Blank		0.000	21:59
Air Blank		0.000	22:01
Subject Sample	# 2	0.000	22:01
Breath Volume		1.828	22:01
Air Blank		0.000	22:02
Control Test		0.081	22:02
Air Blank		0.000	22:03
Diagnostics Check		OK	22:03

Data Download Date/Time: 01/30/2020 14:57

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRELL,CHAD,A
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: DELANEY,TYLER
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 12/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 05:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: HAMPTON

DATE/TIME OF TEST: 12/29/2019 06:23:50

Diagnostics Check		OK	06:21	
Air Blank		0.000	06:21	
Control Test		0.081	06:21	
Air Blank		0.000	06:22	
Subject Sample	# 1		06:22	Subject Test Refused
Breath Volume		0.000	06:22	
Air Blank		0.000	06:22	
Control Test		0.081	06:23	
Air Blank		0.000	06:23	
Diagnostics Check		OK	06:23	

Data Download Date/Time: 01/30/2020 14:57

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORNE,MARSHAWN,L
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: YOUNG,ALEX
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 12/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 05:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:38

SUBJECT'S LAST NAME: HORNE

DATE/TIME OF TEST: 12/29/2019 07:04:45

Diagnostics Check		OK	06:58
Air Blank		0.000	06:59
Control Test		0.081	06:59
Air Blank		0.000	06:59
Subject Sample	# 1	0.211	07:00
Breath Volume		3.839	07:00
Air Blank		0.000	07:00
Air Blank		0.000	07:02
Subject Sample	# 2	0.216	07:03
Breath Volume		3.835	07:03
Air Blank		0.000	07:03
Control Test		0.080	07:04
Air Blank		0.000	07:04
Diagnostics Check		OK	07:04

Data Download Date/Time: 01/30/2020 14:57

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRELL,CHAD,A
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: ADAMS,HOLTON
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 12/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 07:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:10

SUBJECT'S LAST NAME: PODOSOWSKI

DATE/TIME OF TEST: 12/29/2019 08:33:01

Diagnostics Check		OK	08:30	
Air Blank		0.000	08:30	
Control Test		0.081	08:30	
Air Blank		0.000	08:31	
Subject Sample	# 1		08:31	Subject Test Refused
Breath Volume		0.000	08:31	
Air Blank		0.000	08:32	
Control Test		0.081	08:32	
Air Blank		0.000	08:32	
Diagnostics Check		OK	08:33	

Data Download Date/Time: 01/30/2020 14:57

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FURMAN,FREDDIE,R
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: JACKSON,JAMIE
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 12/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 10:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:39

SUBJECT'S LAST NAME: SELF

DATE/TIME OF TEST: 12/31/2019 12:07:19

Diagnostics Check		OK	12:00
Air Blank		0.000	12:00
Control Test		0.081	12:01
Air Blank		0.000	12:01
Subject Sample	# 1	0.241	12:02
Breath Volume		3.449	12:02
Air Blank		0.000	12:03
Air Blank		0.000	12:04
Subject Sample	# 2	0.234	12:05
Breath Volume		1.589	12:05
Air Blank		0.000	12:06
Control Test		0.079	12:06
Air Blank		0.000	12:07
Diagnostics Check		OK	12:07

Data Download Date/Time: 01/30/2020 14:57

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUTTON,KERRY,D
BREATH TEST OPERATOR AGENCY: Lynn Haven PD
ARRESTING OFFICER: DUTTON,KERRY
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 18:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: BRAGG

DATE/TIME OF TEST: 12/05/2019 19:03:39

Diagnostics Check		OK	19:00	
Air Blank		0.000	19:01	
Control Test		0.079	19:01	
Air Blank		0.000	19:02	
Subject Sample	# 1		19:02	Subject Test Refused
Breath Volume		0.000	19:02	
Air Blank		0.000	19:02	
Control Test		0.080	19:03	
Air Blank		0.000	19:03	
Diagnostics Check		OK	19:03	

Data Download Date/Time: 12/23/2019 11:34

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORNE,MARSHAWN,L
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: YOUNG,ALEX
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 12/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG923908
DRY GAS EXPIRY DATE: 08/27/2021
VIOLATION CODE: DUI
ARREST TIME: 05:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:13

SUBJECT'S LAST NAME: HORNE

DATE/TIME OF TEST: 12/29/2019 06:36:56

Diagnostics Check		OK	06:33	
Air Blank		0.000	06:34	
Control Test		0.080	06:34	
Air Blank		0.000	06:35	
Subject Sample	# 1	0.236	06:35	
Breath Volume		2.148	06:35	
Air Blank			06:36	Ambient Fail
Air Blank		0.000	06:36	

Data Download Date/Time: 01/30/2020 14:55

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAENZ,MELISSA,P
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: SAENZ,MELISSA
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 869968
DRY GAS EXPIRY DATE: 12/06/2020
VIOLATION CODE: DUI
ARREST TIME: 19:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: KELLER

DATE/TIME OF TEST: 12/11/2019 20:50:14

Diagnostics Check		OK	20:40
Air Blank		0.000	20:41
Control Test		0.078	20:41
Air Blank		0.000	20:41
Subject Sample	# 1	0.051	20:45
Breath Volume		1.988	20:45
Air Blank		0.000	20:45
Air Blank		0.000	20:47
Subject Sample	# 2	0.049	20:48
Breath Volume		1.769	20:48
Air Blank		0.000	20:49
Control Test		0.080	20:49
Air Blank		0.000	20:50
Diagnostics Check		OK	20:50

Data Download Date/Time: 12/12/2019 16:38

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARRICO,KELLY,D
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: CARRICO,KELLY
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 12/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 869968
DRY GAS EXPIRY DATE: 12/06/2020
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: HURLBIRT

DATE/TIME OF TEST: 12/15/2019 02:24:44

Diagnostics Check		OK	02:16
Air Blank		0.000	02:16
Control Test		0.078	02:17
Air Blank		0.000	02:17
Subject Sample	# 1	0.162	02:19
Breath Volume		1.660	02:19
Air Blank		0.000	02:19
Air Blank		0.000	02:21
Subject Sample	# 2	0.174	02:23
Breath Volume		2.585	02:23
Air Blank		0.000	02:23
Control Test		0.078	02:24
Air Blank		0.000	02:24
Diagnostics Check		OK	02:24

Data Download Date/Time: 01/10/2020 11:16

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICHMOND,BRENDAN,B
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: ELLRICK,JOSH
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 12/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 869968
DRY GAS EXPIRY DATE: 12/06/2020
VIOLATION CODE: DUI
ARREST TIME: 10:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:10

SUBJECT'S LAST NAME: SPANGLER

DATE/TIME OF TEST: 12/30/2019 11:42:12

Diagnostics Check		OK	11:35
Air Blank		0.000	11:35
Control Test		0.077	11:36
Air Blank		0.000	11:36
Subject Sample	# 1	0.000	11:37
Breath Volume		3.214	11:37
Air Blank		0.000	11:38
Air Blank		0.000	11:40
Subject Sample	# 2	0.000	11:40
Breath Volume		3.707	11:40
Air Blank		0.000	11:41
Control Test		0.078	11:41
Air Blank		0.000	11:42
Diagnostics Check		OK	11:42

Data Download Date/Time: 01/10/2020 11:16

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAHAN,WILLIAM,D
BREATH TEST OPERATOR AGENCY: DeFuniak Springs PD
ARRESTING OFFICER: MAHAN,WILLIAM
ARRESTING OFFICER AGENCY: DeFuniak Springs PD
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 17:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:14

SUBJECT'S LAST NAME: GOLIGHTLY

DATE/TIME OF TEST: 12/06/2019 18:49:44

Diagnostics Check		OK	18:43
Air Blank		0.000	18:43
Control Test		0.079	18:44
Air Blank		0.000	18:44
Subject Sample	# 1	0.337	18:45
Breath Volume		3.332	18:45
Air Blank		0.000	18:45
Air Blank		0.000	18:47
Subject Sample	# 2	0.328	18:48
Breath Volume		3.214	18:48
Air Blank		0.000	18:48
Control Test		0.079	18:49
Air Blank		0.000	18:49
Diagnostics Check		OK	18:49

Data Download Date/Time: 12/30/2019 14:43

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAY,JODY,L
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: DAY,JODY
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: MURRY

DATE/TIME OF TEST: 12/19/2019 02:36:58

Diagnostics Check		OK	02:25	
Air Blank		0.000	02:25	
Control Test		0.080	02:25	
Air Blank		0.000	02:26	
Subject Sample	# 1		02:29	No Sample Provided
Breath Volume		0.000	02:29	
Air Blank		0.000	02:29	
Air Blank		0.000	02:31	
Subject Sample	# 2	0.166	02:35	Volume Not Met
Breath Volume		0.390	02:35	
Air Blank		0.000	02:36	
Control Test		0.079	02:36	
Air Blank		0.000	02:36	
Diagnostics Check		OK	02:36	

Data Download Date/Time: 12/30/2019 14:43

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAY,JODY,L
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: DAY,JODY
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: MURRY

DATE/TIME OF TEST: 12/19/2019 02:55:40

Diagnostics Check		OK	02:39	
Air Blank		0.000	02:39	
Control Test		0.079	02:39	
Air Blank		0.000	02:40	
Subject Sample	# 1	0.218	02:42	
Breath Volume		1.582	02:42	
Air Blank		0.000	02:43	
Air Blank		0.000	02:44	
Subject Sample	# 2	0.176	02:47	Volume Not Met
Breath Volume		0.832	02:47	
Air Blank		0.000	02:48	
Air Blank		0.000	02:50	
Subject Sample	# 3	0.153	02:53	Volume Not Met
Breath Volume		0.847	02:53	
Air Blank		0.000	02:54	
Control Test		0.075	02:55	
Air Blank		0.000	02:55	
Diagnostics Check		OK	02:55	

Data Download Date/Time: 12/30/2019 14:43

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAY,JODY,L
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: DAY,JODY
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: MURRY

DATE/TIME OF TEST: 12/19/2019 03:06:43

Diagnostics Check		OK	02:57
Air Blank		0.000	02:58
Control Test		0.078	02:58
Air Blank		0.000	02:59
Subject Sample	# 1	0.194	03:02
Breath Volume		1.355	03:02
Air Blank		0.000	03:02
Air Blank		0.000	03:04
Subject Sample	# 2	0.186	03:05
Breath Volume		1.414	03:05
Air Blank		0.000	03:05
Control Test		0.077	03:06
Air Blank		0.000	03:06
Diagnostics Check		OK	03:06

Data Download Date/Time: 12/30/2019 14:43

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHULTZ,JEREMY,J
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: SCHULTZ,JEREMY
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 12/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:00

SUBJECT'S LAST NAME: MORTON

DATE/TIME OF TEST: 12/30/2019 14:58:06

Diagnostics Check		OK	14:46	
Air Blank		0.000	14:47	
Control Test		0.079	14:47	
Air Blank		0.000	14:47	
Subject Sample	# 1		14:51	No Sample Provided
Breath Volume		0.000	14:51	
Air Blank		0.000	14:51	
Air Blank		0.000	14:53	
Subject Sample	# 2		14:56	No Sample Provided
Breath Volume		0.000	14:56	
Air Blank		0.000	14:57	
Control Test		0.081	14:57	
Air Blank		0.000	14:58	
Diagnostics Check		OK	14:58	

Data Download Date/Time: 01/06/2020 23:35

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHULTZ,JEREMY,J
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: CAISE,JEFFERY
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: MARSH

DATE/TIME OF TEST: 12/07/2019 04:37:03

Diagnostics Check		OK	04:25	
Air Blank		0.000	04:26	
Control Test		0.080	04:26	
Air Blank		0.000	04:26	
Subject Sample	# 1		04:30	No Sample Provided
Breath Volume		0.000	04:30	
Air Blank		0.000	04:30	
Air Blank		0.000	04:32	
Subject Sample	# 2		04:35	No Sample Provided
Breath Volume		0.000	04:35	
Air Blank		0.000	04:36	
Control Test		0.081	04:36	
Air Blank		0.000	04:36	
Diagnostics Check		OK	04:37	

Data Download Date/Time: 12/30/2019 14:53

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMPSON,TREY,A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: SIMPSON,TREY
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: PROB/PAROL
ARREST TIME: 14:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:57

SUBJECT'S LAST NAME: DAVIDSON

DATE/TIME OF TEST: 12/10/2019 15:26:41

Diagnostics Check		OK	15:20
Air Blank		0.000	15:21
Control Test		0.079	15:21
Air Blank		0.000	15:22
Subject Sample	# 1	0.000	15:22
Breath Volume		2.398	15:22
Air Blank		0.000	15:22
Air Blank		0.000	15:24
Subject Sample	# 2	0.000	15:25
Breath Volume		1.914	15:25
Air Blank		0.000	15:25
Control Test		0.078	15:26
Air Blank		0.000	15:26
Diagnostics Check		OK	15:26

Data Download Date/Time: 12/30/2019 14:53

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAYDOS,JONATHAN,A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: MCNAIR,PETER
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: PRICE HORNSBY

DATE/TIME OF TEST: 12/15/2019 05:14:11

Diagnostics Check		OK	05:06
Air Blank		0.000	05:06
Control Test		0.080	05:07
Air Blank		0.000	05:07
Subject Sample	# 1	0.220	05:08
Breath Volume		1.882	05:08
Air Blank		0.000	05:09
Air Blank		0.000	05:11
Subject Sample	# 2	0.225	05:12
Breath Volume		1.917	05:12
Air Blank		0.000	05:13
Control Test		0.081	05:13
Air Blank		0.000	05:14
Diagnostics Check		OK	05:14

Data Download Date/Time: 12/30/2019 14:53

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHULTZ,JEREMY,J
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: COKONOUGHER,ROBERT
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: MCNEIR

DATE/TIME OF TEST: 12/22/2019 00:58:48

Diagnostics Check		OK	00:47	
Air Blank		0.000	00:47	
Control Test		0.080	00:48	
Air Blank		0.000	00:48	
Subject Sample	# 1		00:51	No Sample Provided
Breath Volume		0.000	00:51	
Air Blank		0.000	00:52	
Air Blank		0.000	00:54	
Subject Sample	# 2		00:57	No Sample Provided
Breath Volume		0.000	00:57	
Air Blank		0.000	00:57	
Control Test		0.080	00:58	
Air Blank		0.000	00:58	
Diagnostics Check		OK	00:58	

Data Download Date/Time: 12/30/2019 14:53

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FINK,SHEA,W
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: FINK,SHEA
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 11/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 17:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:23

SUBJECT'S LAST NAME: CARLIN

DATE/TIME OF TEST: 12/01/2019 19:54:48

Diagnostics Check		OK	19:43	
Air Blank		0.000	19:43	
Control Test		0.081	19:44	
Air Blank		0.000	19:44	
Subject Sample	# 1	0.007	19:47	Volume Not Met
Breath Volume		0.050	19:47	
Air Blank		0.000	19:48	
Air Blank		0.000	19:50	
Subject Sample	# 2	0.000	19:53	Volume Not Met
Breath Volume		0.000	19:53	
Air Blank		0.000	19:53	
Control Test		0.079	19:54	
Air Blank		0.000	19:54	
Diagnostics Check		OK	19:54	

Data Download Date/Time: 12/03/2019 10:12

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAJERE,ARIC,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: TEST,TEST
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 11/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: ADMIN
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 12/01/2019 20:04:39

Diagnostics Check		OK	19:58	
Air Blank		0.000	19:59	
Control Test		0.080	19:59	
Air Blank		0.000	20:00	
Subject Sample	# 1	0.000	20:00	
Breath Volume		3.921	20:00	
Air Blank		0.000	20:01	
Air Blank		0.000	20:03	
Subject Sample	# 2		20:03	Subject Test Refused
Breath Volume		0.000	20:03	
Air Blank		0.000	20:03	
Control Test		0.080	20:04	
Air Blank		0.000	20:04	
Diagnostics Check		OK	20:04	

Data Download Date/Time: 12/03/2019 10:12

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FINK,SHEA,W
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: FINK,SHEA
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 11/06/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 17:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: CARLIN

DATE/TIME OF TEST: 12/01/2019 19:56:59

Diagnostics Check		OK	19:50
Air Blank		0.000	19:50
Control Test		0.081	19:51
Air Blank		0.000	19:51
Subject Sample	# 1	0.000	19:52
Breath Volume		1.359	19:52
Air Blank		0.000	19:52
Air Blank		0.000	19:54
Subject Sample	# 2	0.000	19:55
Breath Volume		1.558	19:55
Air Blank		0.000	19:56
Control Test		0.080	19:56
Air Blank		0.000	19:56
Diagnostics Check		OK	19:56

Data Download Date/Time: 12/03/2019 10:12

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,STEVEN,D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK,STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 12/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 21:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: URBINA

DATE/TIME OF TEST: 12/06/2019 22:06:41

Diagnostics Check		OK	22:00
Air Blank		0.000	22:00
Control Test		0.080	22:01
Air Blank		0.000	22:01
Subject Sample	# 1	0.158	22:02
Breath Volume		1.652	22:02
Air Blank		0.000	22:02
Air Blank		0.000	22:04
Subject Sample	# 2	0.156	22:05
Breath Volume		1.750	22:05
Air Blank		0.000	22:05
Control Test		0.080	22:06
Air Blank		0.000	22:06
Diagnostics Check		OK	22:06

Data Download Date/Time: 01/08/2020 11:32

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,STEVEN,D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK,STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 12/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: RACINE

DATE/TIME OF TEST: 12/07/2019 22:51:56

Diagnostics Check		OK	22:45
Air Blank		0.000	22:45
Control Test		0.081	22:46
Air Blank		0.000	22:46
Subject Sample	# 1	0.078	22:47
Breath Volume		2.789	22:47
Air Blank		0.000	22:48
Air Blank		0.000	22:49
Subject Sample	# 2	0.079	22:50
Breath Volume		2.757	22:50
Air Blank		0.000	22:51
Control Test		0.081	22:51
Air Blank		0.000	22:51
Diagnostics Check		OK	22:51

Data Download Date/Time: 01/08/2020 11:32

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NICHOLSON,HORACE,O
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: CAULEY,MICHAEL
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 12/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 19:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:56

SUBJECT'S LAST NAME: DENZIO

DATE/TIME OF TEST: 12/12/2019 20:22:26

Diagnostics Check		OK	20:16
Air Blank		0.000	20:16
Control Test		0.080	20:17
Air Blank		0.000	20:17
Subject Sample	# 1	0.136	20:17
Breath Volume		1.910	20:17
Air Blank		0.000	20:18
Air Blank		0.000	20:20
Subject Sample	# 2	0.146	20:20
Breath Volume		3.585	20:20
Air Blank		0.000	20:21
Control Test		0.080	20:21
Air Blank		0.000	20:22
Diagnostics Check		OK	20:22

Data Download Date/Time: 01/08/2020 11:32

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NICHOLSON,HORACE,O
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: SATALLANTE,DEVON
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 12/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 12/13/2019 00:24:11

Diagnostics Check		OK	00:20	
Air Blank		0.000	00:21	
Control Test		0.081	00:21	
Air Blank		0.000	00:21	
Subject Sample	# 1		00:22	Subject Test Refused
Breath Volume		0.000	00:22	
Air Blank		0.000	00:23	
Control Test		0.080	00:23	
Air Blank		0.000	00:24	
Diagnostics Check		OK	00:24	

Data Download Date/Time: 01/08/2020 11:32

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,STEVEN,D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK,STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 12/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: KELZ

DATE/TIME OF TEST: 12/15/2019 00:36:41

Diagnostics Check		OK	00:30
Air Blank		0.000	00:30
Control Test		0.081	00:31
Air Blank		0.000	00:31
Subject Sample	# 1	0.211	00:32
Breath Volume		2.187	00:32
Air Blank		0.000	00:32
Air Blank		0.000	00:34
Subject Sample	# 2	0.199	00:35
Breath Volume		1.667	00:35
Air Blank		0.000	00:35
Control Test		0.080	00:36
Air Blank		0.000	00:36
Diagnostics Check		OK	00:36

Data Download Date/Time: 01/08/2020 11:32

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,MARK,A
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: 5C20,B
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 12/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 12/16/2019 22:25:25

Diagnostics Check		OK	22:22	
Air Blank		0.000	22:22	
Control Test		0.081	22:23	
Air Blank		0.000	22:23	
Subject Sample	# 1		22:24	Subject Test Refused
Breath Volume		0.000	22:24	
Air Blank		0.000	22:24	
Control Test		0.081	22:24	
Air Blank		0.000	22:25	
Diagnostics Check		OK	22:25	

Data Download Date/Time: 01/08/2020 11:32

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,MARK,A
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: VULETA,B
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 12/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: SNYDER

DATE/TIME OF TEST: 12/16/2019 23:18:23

Diagnostics Check		OK	23:15	
Air Blank		0.000	23:15	
Control Test		0.081	23:16	
Air Blank		0.000	23:16	
Subject Sample	# 1		23:17	Subject Test Refused
Breath Volume		0.000	23:17	
Air Blank		0.000	23:17	
Control Test		0.081	23:17	
Air Blank		0.000	23:18	
Diagnostics Check		OK	23:18	

Data Download Date/Time: 01/08/2020 11:32

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,STEVEN,D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK,STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 12/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 12/22/2019 00:58:16

Diagnostics Check		OK	00:51
Air Blank		0.000	00:51
Control Test		0.080	00:51
Air Blank		0.000	00:52
Subject Sample	# 1	0.204	00:52
Breath Volume		1.820	00:52
Air Blank		0.000	00:53
Air Blank		0.000	00:55
Subject Sample	# 2	0.185	00:56
Breath Volume		1.363	00:56
Air Blank		0.000	00:57
Control Test		0.081	00:57
Air Blank		0.000	00:58
Diagnostics Check		OK	00:58

Data Download Date/Time: 01/08/2020 11:32

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,STEVEN,D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK,STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 12/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 17:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:58

SUBJECT'S LAST NAME: BATOR

DATE/TIME OF TEST: 12/25/2019 18:25:14

Diagnostics Check		OK	18:18
Air Blank		0.000	18:18
Control Test		0.080	18:19
Air Blank		0.000	18:19
Subject Sample	# 1	0.000	18:20
Breath Volume		1.476	18:20
Air Blank		0.000	18:21
Air Blank		0.000	18:23
Subject Sample	# 2	0.000	18:23
Breath Volume		1.429	18:23
Air Blank		0.000	18:24
Control Test		0.081	18:24
Air Blank		0.000	18:25
Diagnostics Check		OK	18:25

Data Download Date/Time: 01/08/2020 11:32

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROMERO,CESAR,
BREATH TEST OPERATOR AGENCY: Okeechobee PD
ARRESTING OFFICER: VULETA,BENJAMIN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 12/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: CHINO MARTINEZ

DATE/TIME OF TEST: 12/26/2019 02:05:19

Diagnostics Check		OK	01:58
Air Blank		0.000	01:58
Control Test		0.081	01:58
Air Blank		0.000	01:59
Subject Sample	# 1	0.190	02:00
Breath Volume		2.179	02:00
Air Blank		0.000	02:01
Air Blank		0.000	02:03
Subject Sample	# 2	0.189	02:03
Breath Volume		1.789	02:03
Air Blank		0.000	02:04
Control Test		0.079	02:04
Air Blank		0.000	02:05
Diagnostics Check		OK	02:05

Data Download Date/Time: 01/08/2020 11:32

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,STEVEN,D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK,STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 12/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: BRYDEBELL

DATE/TIME OF TEST: 12/28/2019 01:13:12

Diagnostics Check		OK	01:05
Air Blank		0.000	01:05
Control Test		0.080	01:05
Air Blank		0.000	01:06
Subject Sample	# 1	0.219	01:07
Breath Volume		3.468	01:07
Air Blank		0.000	01:08
Air Blank		0.000	01:10
Subject Sample	# 2	0.224	01:11
Breath Volume		1.421	01:11
Air Blank		0.000	01:12
Control Test		0.080	01:12
Air Blank		0.000	01:13
Diagnostics Check		OK	01:13

Data Download Date/Time: 01/08/2020 11:32

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,CARL,C
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: VULETA,BENJAMIN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 12/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 12/28/2019 02:54:13

Diagnostics Check		OK	02:50	
Air Blank		0.000	02:51	
Control Test		0.081	02:51	
Air Blank		0.000	02:51	
Subject Sample	# 1		02:52	Subject Test Refused
Breath Volume		0.000	02:52	
Air Blank		0.000	02:53	
Control Test		0.080	02:53	
Air Blank		0.000	02:54	
Diagnostics Check		OK	02:54	

Data Download Date/Time: 01/08/2020 11:32

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,STEVEN,D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK,STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 12/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: CRUZ JR

DATE/TIME OF TEST: 12/28/2019 23:15:46

Diagnostics Check		OK	23:12	
Air Blank		0.000	23:13	
Control Test		0.081	23:13	
Air Blank		0.000	23:14	
Subject Sample	# 1		23:14	Subject Test Refused
Breath Volume		0.000	23:14	
Air Blank		0.000	23:14	
Control Test		0.080	23:15	
Air Blank		0.000	23:15	
Diagnostics Check		OK	23:15	

Data Download Date/Time: 01/08/2020 11:32

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,MARK,A
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: VULETA,B
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 12/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 19:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:24

SUBJECT'S LAST NAME: HARDY

DATE/TIME OF TEST: 12/31/2019 20:47:58

Diagnostics Check		OK	20:44	
Air Blank		0.000	20:45	
Control Test		0.081	20:45	
Air Blank		0.000	20:46	
Subject Sample	# 1		20:46	Subject Test Refused
Breath Volume		0.000	20:46	
Air Blank		0.000	20:47	
Control Test		0.080	20:47	
Air Blank		0.000	20:47	
Diagnostics Check		OK	20:47	

Data Download Date/Time: 01/08/2020 11:32

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,MICHAEL,C
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: WILLIAMS,MICHAEL
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 11/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: BONILLA

DATE/TIME OF TEST: 12/04/2019 21:03:00

Diagnostics Check		OK	20:51	
Air Blank		0.000	20:52	
Control Test		0.079	20:52	
Air Blank		0.000	20:52	
Subject Sample	# 1		20:56	No Sample Provided
Breath Volume		0.000	20:56	
Air Blank		0.000	20:56	
Air Blank		0.000	20:58	
Subject Sample	# 2		21:01	No Sample Provided
Breath Volume		0.000	21:01	
Air Blank		0.000	21:02	
Control Test		0.080	21:02	
Air Blank		0.000	21:02	
Diagnostics Check		OK	21:03	

Data Download Date/Time: 12/30/2019 17:21

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEEFLANG,DANIEL,B
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: WRIGHT,JULIAN
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 11/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: RYAN

DATE/TIME OF TEST: 12/23/2019 22:56:46

Diagnostics Check		OK	22:49
Air Blank		0.000	22:50
Control Test		0.079	22:50
Air Blank		0.000	22:51
Subject Sample	# 1	0.141	22:52
Breath Volume		1.167	22:52
Air Blank		0.000	22:52
Air Blank		0.000	22:54
Subject Sample	# 2	0.136	22:55
Breath Volume		2.554	22:55
Air Blank		0.000	22:55
Control Test		0.077	22:56
Air Blank		0.000	22:56
Diagnostics Check		OK	22:56

Data Download Date/Time: 12/30/2019 17:21

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCINERNEY,JESSE,L
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: THOMAS,BRYCE
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 11/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: LEUNG

DATE/TIME OF TEST: 12/24/2019 23:44:12

Diagnostics Check		OK	23:40	
Air Blank		0.000	23:41	
Control Test		0.080	23:41	
Air Blank		0.000	23:42	
Subject Sample	# 1		23:43	Sequence Aborted
Breath Volume		0.000	23:43	
Air Blank		0.000	23:44	

Data Download Date/Time: 12/30/2019 17:21

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCINERNEY,JESSE,L
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: THOMAS,BRYCE
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 11/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: LEUNG

DATE/TIME OF TEST: 12/24/2019 23:55:13

Diagnostics Check		OK	23:48
Air Blank		0.000	23:49
Control Test		0.079	23:49
Air Blank		0.000	23:49
Subject Sample	# 1	0.186	23:50
Breath Volume		2.335	23:50
Air Blank		0.000	23:51
Air Blank		0.000	23:53
Subject Sample	# 2	0.191	23:53
Breath Volume		2.628	23:53
Air Blank		0.000	23:54
Control Test		0.078	23:54
Air Blank		0.000	23:55
Diagnostics Check		OK	23:55

Data Download Date/Time: 12/30/2019 17:21

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,SALVADOR,
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: BATTISTELLI,TAYLOR
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 11/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 11:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:30

SUBJECT'S LAST NAME: BENITEZ

DATE/TIME OF TEST: 12/25/2019 11:56:17

Diagnostics Check		OK	11:50
Air Blank		0.000	11:50
Control Test		0.080	11:50
Air Blank		0.000	11:51
Subject Sample	# 1	0.296	11:51
Breath Volume		3.699	11:51
Air Blank		0.000	11:52
Air Blank		0.000	11:54
Subject Sample	# 2	0.303	11:54
Breath Volume		2.531	11:54
Air Blank		0.000	11:55
Control Test		0.079	11:55
Air Blank		0.000	11:56
Diagnostics Check		OK	11:56

Data Download Date/Time: 12/30/2019 17:21

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,MICHAEL,C
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: GASPAR,BENJAMIN
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 11/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: DECIMUS

DATE/TIME OF TEST: 12/27/2019 00:01:55

Diagnostics Check		OK	23:55
Air Blank		0.000	23:55
Control Test		0.079	23:56
Air Blank		0.000	23:56
Subject Sample	# 2	0.152	23:57
Breath Volume		3.050	23:57
Air Blank		0.000	23:58
Air Blank		0.000	23:59
Subject Sample	# 1	0.155	00:00
Breath Volume		2.921	00:00
Air Blank		0.000	00:01
Control Test		0.078	00:01
Air Blank		0.000	00:01
Diagnostics Check		OK	00:01

Data Download Date/Time: 12/30/2019 17:21

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCINERNEY,JESSE,L
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: KLECKLEY,LAQUESHA
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 11/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: DELGADO

DATE/TIME OF TEST: 12/29/2019 03:37:00

Diagnostics Check		OK	03:30
Air Blank		0.000	03:30
Control Test		0.079	03:31
Air Blank		0.000	03:31
Subject Sample	# 1	0.181	03:32
Breath Volume		3.167	03:32
Air Blank		0.000	03:33
Air Blank		0.000	03:35
Subject Sample	# 2	0.178	03:35
Breath Volume		3.410	03:35
Air Blank		0.000	03:36
Control Test		0.077	03:36
Air Blank		0.000	03:36
Diagnostics Check		OK	03:37

Data Download Date/Time: 12/30/2019 17:21

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIGNORELLI,A,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZE,B
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: RODGERS

DATE/TIME OF TEST: 12/01/2019 22:58:18

Diagnostics Check		OK	22:50
Air Blank		0.000	22:51
Control Test		0.080	22:51
Air Blank		0.000	22:51
Subject Sample	# 1	0.133	22:52
Breath Volume		1.597	22:52
Air Blank		0.000	22:53
Air Blank		0.000	22:55
Subject Sample	# 2	0.131	22:56
Breath Volume		1.554	22:56
Air Blank		0.000	22:57
Control Test		0.079	22:57
Air Blank		0.000	22:58
Diagnostics Check		OK	22:58

Data Download Date/Time: 12/13/2019 06:14

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIGNORELLI,A,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: RUIZ,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: STRACKE

DATE/TIME OF TEST: 12/01/2019 23:41:41

Diagnostics Check		OK	23:38	
Air Blank		0.000	23:38	
Control Test		0.079	23:39	
Air Blank		0.000	23:39	
Subject Sample	# 1		23:40	Subject Test Refused
Breath Volume		0.000	23:40	
Air Blank		0.000	23:40	
Control Test		0.080	23:41	
Air Blank		0.000	23:41	
Diagnostics Check		OK	23:41	

Data Download Date/Time: 12/13/2019 06:14

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANDY,KEVIN,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,BRANDON
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: HENDRICKS

DATE/TIME OF TEST: 12/02/2019 03:02:36

Diagnostics Check		OK	02:56
Air Blank		0.000	02:56
Control Test		0.080	02:56
Air Blank		0.000	02:57
Subject Sample	# 1	0.128	02:58
Breath Volume		3.546	02:58
Air Blank		0.000	02:58
Air Blank		0.000	03:00
Subject Sample	# 2	0.124	03:01
Breath Volume		2.968	03:01
Air Blank		0.000	03:01
Control Test		0.080	03:02
Air Blank		0.000	03:02
Diagnostics Check		OK	03:02

Data Download Date/Time: 12/13/2019 06:14

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DEPUTY,T
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: FLORES,DEPUTY
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 12/03/2019 22:38:02

Diagnostics Check		OK	22:31
Air Blank		0.000	22:31
Control Test		0.080	22:32
Air Blank		0.000	22:32
Subject Sample	# 1	0.128	22:33
Breath Volume		3.074	22:33
Air Blank		0.000	22:34
Air Blank		0.000	22:35
Subject Sample	# 2	0.134	22:36
Breath Volume		2.433	22:36
Air Blank		0.000	22:37
Control Test		0.081	22:37
Air Blank		0.000	22:37
Diagnostics Check		OK	22:38

Data Download Date/Time: 12/13/2019 06:14

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEARSE,ANGIE,D
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: SAINT ELOI,A
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: NEIDECKER

DATE/TIME OF TEST: 12/04/2019 01:27:34

Diagnostics Check		OK	01:24	
Air Blank		0.000	01:25	
Control Test		0.081	01:25	
Air Blank		0.000	01:25	
Subject Sample	# 1		01:26	Subject Test Refused
Breath Volume		0.000	01:26	
Air Blank		0.000	01:26	
Control Test		0.080	01:26	
Air Blank		0.000	01:27	
Diagnostics Check		OK	01:27	

Data Download Date/Time: 12/13/2019 06:14

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FONSECA,GIOVANNY,
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: FONSECA,GIOVANNY
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: EBERLY

DATE/TIME OF TEST: 12/05/2019 03:53:00

Diagnostics Check		OK	03:49	
Air Blank		0.000	03:50	
Control Test		0.080	03:50	
Air Blank		0.000	03:51	
Subject Sample	# 1		03:51	Subject Test Refused
Breath Volume		0.000	03:51	
Air Blank		0.000	03:52	
Control Test		0.080	03:52	
Air Blank		0.000	03:52	
Diagnostics Check		OK	03:53	

Data Download Date/Time: 12/13/2019 06:14

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIGNORELLI,A,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,B
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 21:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:44

SUBJECT'S LAST NAME: UNDERWOD

DATE/TIME OF TEST: 12/05/2019 22:13:04

Diagnostics Check		OK	22:06
Air Blank		0.000	22:06
Control Test		0.081	22:07
Air Blank		0.000	22:07
Subject Sample	# 1	0.035	22:08
Breath Volume		1.976	22:08
Air Blank		0.000	22:09
Air Blank		0.000	22:10
Subject Sample	# 2	0.035	22:11
Breath Volume		2.070	22:11
Air Blank		0.000	22:12
Control Test		0.081	22:12
Air Blank		0.000	22:12
Diagnostics Check		OK	22:13

Data Download Date/Time: 12/13/2019 06:14

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANDY',KEVIN,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,BRANDON
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: BEACH

DATE/TIME OF TEST: 12/06/2019 02:37:04

Diagnostics Check		OK	02:33	
Air Blank		0.000	02:34	
Control Test		0.081	02:34	
Air Blank		0.000	02:35	
Subject Sample	# 1		02:35	Subject Test Refused
Breath Volume		0.000	02:35	
Air Blank		0.000	02:36	
Control Test		0.080	02:36	
Air Blank		0.000	02:36	
Diagnostics Check		OK	02:37	

Data Download Date/Time: 12/13/2019 06:14

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANDY,KEVIN,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MANGEL,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 12/06/2019 03:04:28

Diagnostics Check		OK	02:57
Air Blank		0.000	02:58
Control Test		0.080	02:58
Air Blank		0.000	02:58
Subject Sample	# 1	0.144	02:59
Breath Volume		1.246	02:59
Air Blank		0.000	03:00
Air Blank		0.000	03:01
Subject Sample	# 2	0.148	03:02
Breath Volume		1.687	03:02
Air Blank		0.000	03:03
Control Test		0.080	03:03
Air Blank		0.000	03:04
Diagnostics Check		OK	03:04

Data Download Date/Time: 12/13/2019 06:14

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MAEGAN,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: JAWORSKI,K
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 16:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:25

SUBJECT'S LAST NAME: MAURO

DATE/TIME OF TEST: 12/06/2019 18:13:59

Diagnostics Check		OK	18:08
Air Blank		0.000	18:08
Control Test		0.081	18:08
Air Blank		0.000	18:09
Subject Sample	# 1	0.000	18:09
Breath Volume		1.550	18:09
Air Blank		0.000	18:10
Air Blank		0.000	18:12
Subject Sample	# 2	0.000	18:12
Breath Volume		1.554	18:12
Air Blank		0.000	18:13
Control Test		0.080	18:13
Air Blank		0.000	18:13
Diagnostics Check		OK	18:13

Data Download Date/Time: 12/13/2019 06:14

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AYALA,KASSANDRA,
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: AYALA,KASSANDRA
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: SOARE

DATE/TIME OF TEST: 12/07/2019 02:48:03

Diagnostics Check		OK	02:40
Air Blank		0.000	02:40
Control Test		0.080	02:41
Air Blank		0.000	02:41
Subject Sample	# 1	0.195	02:42
Breath Volume		1.851	02:42
Air Blank		0.000	02:43
Air Blank		0.000	02:45
Subject Sample	# 2	0.197	02:46
Breath Volume		1.136	02:46
Air Blank		0.000	02:47
Control Test		0.079	02:47
Air Blank		0.000	02:47
Diagnostics Check		OK	02:48

Data Download Date/Time: 12/13/2019 06:14

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KING,SHAKAYLA,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MANGEL,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 12/07/2019 03:13:23

Diagnostics Check		OK	03:10	
Air Blank		0.000	03:10	
Control Test		0.080	03:10	
Air Blank		0.000	03:11	
Subject Sample	# 1		03:12	Subject Test Refused
Breath Volume		0.000	03:12	
Air Blank		0.000	03:12	
Control Test		0.080	03:12	
Air Blank		0.000	03:13	
Diagnostics Check		OK	03:13	

Data Download Date/Time: 12/13/2019 06:14

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KUEHN,KYLE,B
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: KUEHN,KYLE
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 19:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:18

SUBJECT'S LAST NAME: HORSKIN

DATE/TIME OF TEST: 12/07/2019 20:48:49

Diagnostics Check		OK	20:38	
Air Blank		0.000	20:38	
Control Test		0.080	20:39	
Air Blank		0.000	20:39	
Subject Sample	# 1	0.244	20:42	Volume Not Met
Breath Volume		0.289	20:42	
Air Blank		0.000	20:43	
Air Blank		0.000	20:45	
Subject Sample	# 2	0.265	20:45	
Breath Volume		1.410	20:45	
Air Blank		0.000	20:46	
Air Blank		0.000	20:48	
Subject Sample	# 3	0.258	20:48	
Breath Volume		1.800	20:48	
Air Blank			20:48	RFI Detect
Air Blank			20:48	RFI Detect

Data Download Date/Time: 12/13/2019 06:14

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MAEGAN,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MANGEL,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: RICHESIN

DATE/TIME OF TEST: 12/07/2019 22:59:40

Diagnostics Check		OK	22:53
Air Blank		0.000	22:54
Control Test		0.080	22:54
Air Blank		0.000	22:54
Subject Sample	# 1	0.151	22:55
Breath Volume		1.539	22:55
Air Blank		0.000	22:55
Air Blank		0.000	22:57
Subject Sample	# 2	0.153	22:58
Breath Volume		1.585	22:58
Air Blank		0.000	22:58
Control Test		0.079	22:59
Air Blank		0.000	22:59
Diagnostics Check		OK	22:59

Data Download Date/Time: 12/13/2019 06:14

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIGNORELLI,A,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: BRANDES,M
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 19:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:51

SUBJECT'S LAST NAME: AMELIA

DATE/TIME OF TEST: 12/08/2019 20:28:43

Diagnostics Check		OK	20:20
Air Blank		0.000	20:21
Control Test		0.080	20:21
Air Blank		0.000	20:21
Subject Sample	# 1	0.173	20:22
Breath Volume		1.816	20:22
Air Blank		0.000	20:23
Air Blank		0.000	20:25
Subject Sample	# 2	0.160	20:27
Breath Volume		1.816	20:27
Air Blank		0.000	20:27
Control Test		0.080	20:28
Air Blank		0.000	20:28
Diagnostics Check		OK	20:28

Data Download Date/Time: 12/13/2019 06:14

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WLLIAMS,DEPUTY,T
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZE,OFFICER
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: LYON

DATE/TIME OF TEST: 12/10/2019 22:50:27

Diagnostics Check		OK	22:48	
Air Blank		0.000	22:49	
Control Test		0.081	22:49	
Air Blank		0.000	22:49	
Subject Sample	# 1		22:50	RFI Detect
Breath Volume		0.000	22:50	
Air Blank			22:50	RFI Detect
Air Blank			22:50	RFI Detect

Data Download Date/Time: 12/13/2019 06:14

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DEPUTY,T
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZE,OFFICER
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 11/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: LYON

DATE/TIME OF TEST: 12/10/2019 22:59:11

Diagnostics Check		OK	22:52
Air Blank		0.000	22:53
Control Test		0.080	22:53
Air Blank		0.000	22:54
Subject Sample	# 1	0.000	22:54
Breath Volume		1.789	22:54
Air Blank		0.000	22:55
Air Blank		0.000	22:57
Subject Sample	# 2	0.000	22:57
Breath Volume		1.503	22:57
Air Blank		0.000	22:58
Control Test		0.080	22:58
Air Blank		0.000	22:59
Diagnostics Check		OK	22:59

Data Download Date/Time: 12/13/2019 06:14

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YORK,JIMMY,A
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: NIEVES,LUIS
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 13:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:59

SUBJECT'S LAST NAME: HERNANDEZ FLORES

DATE/TIME OF TEST: 12/13/2019 14:26:50

Diagnostics Check		OK	14:19
Air Blank		0.000	14:20
Control Test		0.080	14:20
Air Blank		0.000	14:21
Subject Sample	# 1	0.201	14:21
Breath Volume		1.351	14:21
Air Blank		0.000	14:22
Air Blank		0.000	14:24
Subject Sample	# 2	0.201	14:25
Breath Volume		1.175	14:25
Air Blank		0.000	14:25
Control Test		0.080	14:26
Air Blank		0.000	14:26
Diagnostics Check		OK	14:26

Data Download Date/Time: 01/03/2020 05:43

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHILE,C,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,B
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 12/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: MCISAAC

DATE/TIME OF TEST: 12/13/2019 22:50:58

Diagnostics Check		OK	22:48	
Air Blank		0.000	22:48	
Control Test		0.080	22:48	
Air Blank		0.000	22:49	
Subject Sample	# 1		22:49	Subject Test Refused
Breath Volume		0.000	22:49	
Air Blank		0.000	22:50	
Control Test		0.080	22:50	
Air Blank		0.000	22:50	
Diagnostics Check		OK	22:50	

Data Download Date/Time: 01/03/2020 05:43

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANDY,KEVIN,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,BRANDON
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 12/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: SHAULIS

DATE/TIME OF TEST: 12/14/2019 02:33:54

Diagnostics Check		OK	02:31	
Air Blank		0.000	02:31	
Control Test		0.080	02:31	
Air Blank		0.000	02:32	
Subject Sample	# 1		02:32	Subject Test Refused
Breath Volume		0.000	02:32	
Air Blank		0.000	02:33	
Control Test		0.080	02:33	
Air Blank		0.000	02:33	
Diagnostics Check		OK	02:33	

Data Download Date/Time: 01/03/2020 05:43

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DEPUTY,T
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,OFFICER
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 12/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: HODGES

DATE/TIME OF TEST: 12/14/2019 23:37:53

Diagnostics Check		OK	23:35	
Air Blank		0.000	23:35	
Control Test		0.080	23:35	
Air Blank		0.000	23:36	
Subject Sample	# 1		23:36	Subject Test Refused
Breath Volume		0.000	23:36	
Air Blank		0.000	23:36	
Control Test		0.080	23:37	
Air Blank		0.000	23:37	
Diagnostics Check		OK	23:37	

Data Download Date/Time: 01/03/2020 05:43

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MAEGAN,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: RUIZ,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: CEJKA

DATE/TIME OF TEST: 12/14/2019 23:57:43

Diagnostics Check		OK	23:52	
Air Blank		0.000	23:52	
Control Test		0.080	23:52	
Air Blank		0.000	23:53	
Subject Sample	# 1	0.163	23:53	
Breath Volume		1.871	23:53	
Air Blank		0.000	23:54	
Air Blank		0.000	23:56	
Subject Sample	# 2	0.164	23:56	
Breath Volume		1.488	23:56	
Air Blank			23:57	RFI Detect
Air Blank		0.000	23:57	

Data Download Date/Time: 01/03/2020 05:43

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KANTORSKI,MARK,A
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: INGRAM,JENNIFER
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 12/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: GOLLINGE

DATE/TIME OF TEST: 12/15/2019 02:19:05

Diagnostics Check		OK	02:12
Air Blank		0.000	02:12
Control Test		0.080	02:12
Air Blank		0.000	02:13
Subject Sample	# 1	0.148	02:14
Breath Volume		1.902	02:14
Air Blank		0.000	02:14
Air Blank		0.000	02:16
Subject Sample	# 2	0.146	02:17
Breath Volume		1.835	02:17
Air Blank		0.000	02:18
Control Test		0.081	02:18
Air Blank		0.000	02:19
Diagnostics Check		OK	02:19

Data Download Date/Time: 01/03/2020 05:43

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,DEVIN,C
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MANGEL,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: TROUTMAN

DATE/TIME OF TEST: 12/15/2019 02:46:58

Diagnostics Check		OK	02:44	
Air Blank		0.000	02:44	
Control Test		0.080	02:44	
Air Blank		0.000	02:45	
Subject Sample	# 1		02:45	Subject Test Refused
Breath Volume		0.000	02:45	
Air Blank		0.000	02:46	
Control Test		0.080	02:46	
Air Blank		0.000	02:46	
Diagnostics Check		OK	02:46	

Data Download Date/Time: 01/03/2020 05:43

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANDY,KEVIN,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,BRANDON
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 12/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: HOPE

DATE/TIME OF TEST: 12/15/2019 05:06:12

Diagnostics Check		OK	04:59
Air Blank		0.000	04:59
Control Test		0.080	05:00
Air Blank		0.000	05:00
Subject Sample	# 1	0.000	05:01
Breath Volume		3.414	05:01
Air Blank		0.000	05:02
Air Blank		0.000	05:04
Subject Sample	# 2	0.000	05:04
Breath Volume		3.542	05:04
Air Blank		0.000	05:05
Control Test		0.081	05:05
Air Blank		0.000	05:06
Diagnostics Check		OK	05:06

Data Download Date/Time: 01/03/2020 05:44

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MAEGAN,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: FRICKE,J
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 12/15/2019 21:52:34

Diagnostics Check		OK	21:49	
Air Blank		0.000	21:50	
Control Test		0.081	21:50	
Air Blank		0.000	21:51	
Subject Sample	# 1		21:51	Subject Test Refused
Breath Volume		0.000	21:51	
Air Blank		0.000	21:51	
Control Test		0.081	21:51	
Air Blank		0.000	21:52	
Diagnostics Check		OK	21:52	

Data Download Date/Time: 01/03/2020 05:44

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEARSE,ANGIE,D
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,BRANDON
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 12/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 12/16/2019 00:47:36

Diagnostics Check		OK	00:41
Air Blank		0.000	00:41
Control Test		0.082	00:41
Air Blank		0.000	00:42
Subject Sample	# 1	0.112	00:42
Breath Volume		2.574	00:42
Air Blank		0.000	00:43
Air Blank		0.000	00:45
Subject Sample	# 2	0.106	00:45
Breath Volume		1.421	00:45
Air Blank		0.000	00:46
Control Test		0.082	00:46
Air Blank		0.000	00:47
Diagnostics Check		OK	00:47

Data Download Date/Time: 01/03/2020 05:19

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANDY,KEVIN,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,BRANDON
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 12/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 12/16/2019 03:12:02

Diagnostics Check		OK	03:09	
Air Blank		0.000	03:09	
Control Test		0.082	03:09	
Air Blank		0.000	03:10	
Subject Sample	# 1		03:10	Subject Test Refused
Breath Volume		0.000	03:10	
Air Blank		0.000	03:11	
Control Test		0.082	03:11	
Air Blank		0.000	03:11	
Diagnostics Check		OK	03:12	

Data Download Date/Time: 01/03/2020 05:19

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MAEGAN,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: RUIZ,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 19:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: PRICE

DATE/TIME OF TEST: 12/16/2019 21:42:05

Diagnostics Check		OK	21:31	
Air Blank		0.000	21:32	
Control Test		0.082	21:32	
Air Blank		0.000	21:33	
Subject Sample	# 1	0.233	21:36	Volume Not Met
Breath Volume		0.015	21:36	
Air Blank		0.000	21:37	
Air Blank		0.000	21:38	
Subject Sample	# 2		21:40	Subject Test Refused
Breath Volume		0.000	21:40	
Air Blank		0.000	21:41	
Control Test		0.080	21:41	
Air Blank		0.000	21:42	
Diagnostics Check		OK	21:42	

Data Download Date/Time: 01/03/2020 05:19

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIGNORELLI,A,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: RISPOLI,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: WIMETT

DATE/TIME OF TEST: 12/18/2019 23:26:52

Diagnostics Check		OK	23:19
Air Blank		0.000	23:20
Control Test		0.083	23:20
Air Blank		0.000	23:21
Subject Sample	# 1	0.199	23:21
Breath Volume		1.417	23:21
Air Blank		0.000	23:22
Air Blank		0.000	23:24
Subject Sample	# 2	0.199	23:25
Breath Volume		1.882	23:25
Air Blank		0.000	23:25
Control Test		0.082	23:26
Air Blank		0.000	23:26
Diagnostics Check		OK	23:26

Data Download Date/Time: 01/03/2020 05:19

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIGNORELLI,A,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,B
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 12/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: WINTERLYNN

DATE/TIME OF TEST: 12/19/2019 00:00:23

Diagnostics Check		OK	23:53
Air Blank		0.000	23:54
Control Test		0.082	23:54
Air Blank		0.000	23:55
Subject Sample	# 1	0.000	23:55
Breath Volume		1.195	23:55
Air Blank		0.000	23:56
Air Blank		0.000	23:58
Subject Sample	# 2	0.000	23:59
Breath Volume		2.128	23:59
Air Blank		0.000	23:59
Control Test		0.082	23:59
Air Blank		0.000	00:00
Diagnostics Check		OK	00:00

Data Download Date/Time: 01/03/2020 05:19

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,CODY,L
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: BROWN,CODY
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 12/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: MEIRA

DATE/TIME OF TEST: 12/20/2019 03:46:38

Diagnostics Check		OK	03:40
Air Blank		0.000	03:40
Control Test		0.083	03:40
Air Blank		0.000	03:41
Subject Sample	# 1	0.175	03:41
Breath Volume		1.445	03:41
Air Blank		0.000	03:42
Air Blank		0.000	03:44
Subject Sample	# 2	0.188	03:45
Breath Volume		2.984	03:45
Air Blank		0.000	03:45
Control Test		0.082	03:46
Air Blank		0.000	03:46
Diagnostics Check		OK	03:46

Data Download Date/Time: 01/03/2020 05:19

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RICKY,
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: PEREZ,RICKY
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 12/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 16:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: KEATING

DATE/TIME OF TEST: 12/20/2019 17:31:00

Diagnostics Check		OK	17:22	
Air Blank		0.000	17:23	
Control Test		0.083	17:23	
Air Blank		0.000	17:23	
Subject Sample	# 1		17:26	No Sample Provided
Breath Volume		0.000	17:26	
Air Blank		0.000	17:27	
Air Blank		0.000	17:29	
Subject Sample	# 2		17:29	Subject Test Refused
Breath Volume		0.000	17:29	
Air Blank		0.000	17:30	
Control Test		0.083	17:30	
Air Blank		0.000	17:30	
Diagnostics Check		OK	17:31	

Data Download Date/Time: 01/03/2020 05:19

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARROQUIN,RICHARD,J
BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)
ARRESTING OFFICER: MARROQUIN,RICHARD
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 12/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: JORGE

DATE/TIME OF TEST: 12/22/2019 03:30:08

Diagnostics Check		OK	03:27	
Air Blank		0.000	03:27	
Control Test		0.083	03:28	
Air Blank		0.000	03:28	
Subject Sample	# 1		03:28	Subject Test Refused
Breath Volume		0.000	03:28	
Air Blank		0.000	03:29	
Control Test		0.083	03:29	
Air Blank		0.000	03:30	
Diagnostics Check		OK	03:30	

Data Download Date/Time: 01/03/2020 05:19

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DEPUTY,T
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: FLORES,DEPUTY
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: THORNE

DATE/TIME OF TEST: 12/22/2019 22:00:24

Diagnostics Check		OK	21:52
Air Blank		0.000	21:53
Control Test		0.083	21:53
Air Blank		0.000	21:54
Subject Sample	# 1	0.309	21:55
Breath Volume		1.390	21:55
Air Blank		0.000	21:56
Air Blank		0.000	21:57
Subject Sample	# 2	0.320	21:58
Breath Volume		2.429	21:58
Air Blank		0.000	21:59
Control Test		0.082	21:59
Air Blank		0.000	22:00
Diagnostics Check		OK	22:00

Data Download Date/Time: 01/03/2020 05:19

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEARSE,ANGIE,D
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: FERGUSON,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: LEON

DATE/TIME OF TEST: 12/23/2019 00:52:03

Diagnostics Check		OK	00:45
Air Blank		0.000	00:46
Control Test		0.083	00:46
Air Blank		0.000	00:47
Subject Sample	# 1	0.000	00:47
Breath Volume		1.867	00:47
Air Blank		0.000	00:48
Air Blank		0.000	00:50
Subject Sample	# 2	0.000	00:50
Breath Volume		1.156	00:50
Air Blank		0.000	00:51
Control Test		0.083	00:51
Air Blank		0.000	00:51
Diagnostics Check		OK	00:52

Data Download Date/Time: 01/03/2020 05:19

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAYMOND,FRANK,W
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: SHAW,N
ARRESTING OFFICER AGENCY: Indian River Shores
LAST AGENCY INSPECTION: 12/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: ZINNA

DATE/TIME OF TEST: 12/25/2019 01:00:21

Diagnostics Check		OK	00:56	
Air Blank		0.000	00:57	
Control Test		0.082	00:57	
Air Blank		0.000	00:57	
Subject Sample	# 1		00:58	Subject Test Refused
Breath Volume		0.000	00:58	
Air Blank		0.000	00:59	
Control Test		0.083	00:59	
Air Blank		0.000	01:00	
Diagnostics Check		OK	01:00	

Data Download Date/Time: 01/03/2020 05:19

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIGNORELLI,A,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MANGEL,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: STEPANEK

DATE/TIME OF TEST: 12/25/2019 23:23:07

Diagnostics Check		OK	23:16
Air Blank		0.000	23:16
Control Test		0.084	23:17
Air Blank		0.000	23:17
Subject Sample	# 1	0.054	23:18
Breath Volume		1.667	23:18
Air Blank		0.000	23:19
Air Blank		0.000	23:20
Subject Sample	# 2	0.052	23:21
Breath Volume		1.808	23:21
Air Blank		0.000	23:22
Control Test		0.083	23:22
Air Blank		0.000	23:23
Diagnostics Check		OK	23:23

Data Download Date/Time: 01/03/2020 05:19

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YORK,JIMMY,A
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: PIANELLI,VICTORIA
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 11:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:29

SUBJECT'S LAST NAME: GARCIA VILLALPA

DATE/TIME OF TEST: 12/26/2019 13:57:22

Diagnostics Check		OK	13:50
Air Blank		0.000	13:51
Control Test		0.083	13:51
Air Blank		0.000	13:52
Subject Sample	# 1	0.169	13:52
Breath Volume		3.066	13:52
Air Blank		0.000	13:53
Air Blank		0.000	13:55
Subject Sample	# 2	0.167	13:55
Breath Volume		3.117	13:55
Air Blank		0.000	13:56
Control Test		0.082	13:56
Air Blank		0.000	13:57
Diagnostics Check		OK	13:57

Data Download Date/Time: 01/03/2020 05:19

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIGNORELLI,A,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MANGEL,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 12/26/2019 22:35:32

Diagnostics Check		OK	22:27
Air Blank		0.000	22:27
Control Test		0.083	22:28
Air Blank		0.000	22:28
Subject Sample	# 1	0.181	22:30
Breath Volume		1.238	22:30
Air Blank		0.000	22:31
Air Blank		0.000	22:32
Subject Sample	# 2	0.196	22:33
Breath Volume		1.410	22:33
Air Blank		0.000	22:34
Control Test		0.082	22:34
Air Blank		0.000	22:35
Diagnostics Check		OK	22:35

Data Download Date/Time: 01/03/2020 05:19

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANDY,KEVIN,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MANGEL,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: LIRAMEAU

DATE/TIME OF TEST: 12/28/2019 02:53:13

Diagnostics Check		OK	02:46
Air Blank		0.000	02:46
Control Test		0.083	02:47
Air Blank		0.000	02:47
Subject Sample	# 1	0.246	02:48
Breath Volume		2.011	02:48
Air Blank		0.000	02:49
Air Blank		0.000	02:51
Subject Sample	# 2	0.244	02:51
Breath Volume		3.015	02:51
Air Blank		0.000	02:52
Control Test		0.082	02:52
Air Blank		0.000	02:53
Diagnostics Check		OK	02:53

Data Download Date/Time: 01/03/2020 05:19

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KANTORSKI,MARK,A
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: FINNEGAN,SERGEANT
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 05:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:45

SUBJECT'S LAST NAME: PERRY

DATE/TIME OF TEST: 12/28/2019 06:19:02

Diagnostics Check		OK	06:11
Air Blank		0.000	06:12
Control Test		0.083	06:12
Air Blank		0.000	06:13
Subject Sample	# 1	0.000	06:14
Breath Volume		1.695	06:14
Air Blank		0.000	06:14
Air Blank		0.000	06:16
Subject Sample	# 2	0.000	06:17
Breath Volume		1.921	06:17
Air Blank		0.000	06:18
Control Test		0.082	06:18
Air Blank		0.000	06:18
Diagnostics Check		OK	06:19

Data Download Date/Time: 01/03/2020 05:19

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIST,SHARON,D
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: RADOCK,M
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 12/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 04:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:00

SUBJECT'S LAST NAME: P615777984160

DATE/TIME OF TEST: 12/28/2019 10:32:14

Diagnostics Check		OK	10:25
Air Blank		0.000	10:25
Control Test		0.083	10:25
Air Blank		0.000	10:26
Subject Sample	# 1	0.000	10:27
Breath Volume		3.089	10:27
Air Blank		0.000	10:27
Air Blank		0.000	10:29
Subject Sample	# 2	0.000	10:30
Breath Volume		3.523	10:30
Air Blank		0.000	10:31
Control Test		0.083	10:31
Air Blank		0.000	10:32
Diagnostics Check		OK	10:32

Data Download Date/Time: 01/03/2020 05:19

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIGNORELLI,A,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: ORCHOSTLEY,G
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 15:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:15

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 12/28/2019 16:45:05

Diagnostics Check		OK	16:37
Air Blank		0.000	16:38
Control Test		0.082	16:38
Air Blank		0.000	16:39
Subject Sample	# 1	0.333	16:40
Breath Volume		1.632	16:40
Air Blank		0.000	16:40
Air Blank		0.000	16:42
Subject Sample	# 2	0.335	16:43
Breath Volume		1.898	16:43
Air Blank		0.000	16:44
Control Test		0.081	16:44
Air Blank		0.000	16:45
Diagnostics Check		OK	16:45

Data Download Date/Time: 01/03/2020 05:19

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KANTORSKI,MARK,A
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: RUIZ,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: GIERSDORF

DATE/TIME OF TEST: 12/29/2019 01:26:58

Diagnostics Check		OK	01:24	
Air Blank		0.000	01:24	
Control Test		0.082	01:24	
Air Blank		0.000	01:25	
Subject Sample	# 1		01:25	Subject Test Refused
Breath Volume		0.000	01:25	
Air Blank		0.000	01:26	
Control Test		0.083	01:26	
Air Blank		0.000	01:26	
Diagnostics Check		OK	01:26	

Data Download Date/Time: 01/03/2020 05:19

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DEPUTY,T
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MACIAS,DEPUTY
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 19:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: DE FRANCO

DATE/TIME OF TEST: 12/29/2019 20:29:18

Diagnostics Check		OK	20:22
Air Blank		0.000	20:23
Control Test		0.083	20:23
Air Blank		0.000	20:24
Subject Sample	# 1	0.176	20:24
Breath Volume		1.226	20:24
Air Blank		0.000	20:25
Air Blank		0.000	20:27
Subject Sample	# 2	0.185	20:27
Breath Volume		1.109	20:27
Air Blank		0.000	20:28
Control Test		0.082	20:28
Air Blank		0.000	20:29
Diagnostics Check		OK	20:29

Data Download Date/Time: 01/03/2020 05:19

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAYMOND,FRANK,W
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: INGRAM,J
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 12/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: ZAMUDIO

DATE/TIME OF TEST: 12/30/2019 00:50:54

Diagnostics Check		OK	00:47	
Air Blank		0.000	00:48	
Control Test		0.083	00:48	
Air Blank		0.000	00:49	
Subject Sample	# 1		00:49	Subject Test Refused
Breath Volume		0.000	00:49	
Air Blank		0.000	00:49	
Control Test		0.083	00:50	
Air Blank		0.000	00:50	
Diagnostics Check		OK	00:50	

Data Download Date/Time: 01/03/2020 05:19

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DEPUTY,.
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,OFFICER
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 12/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 18:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: CASSEUS

DATE/TIME OF TEST: 12/31/2019 19:45:53

Diagnostics Check		OK	19:31	
Air Blank		0.000	19:32	
Control Test		0.082	19:32	
Air Blank		0.000	19:32	
Subject Sample	# 1	0.000	19:36	Volume Not Met
Breath Volume		1.222	19:36	
Air Blank		0.000	19:36	
Air Blank		0.000	19:38	
Subject Sample	# 2	0.000	19:40	
Breath Volume		1.390	19:40	
Air Blank		0.000	19:41	
Air Blank		0.000	19:43	
Subject Sample	# 3	0.000	19:44	
Breath Volume		1.281	19:44	
Air Blank		0.000	19:44	
Control Test		0.081	19:45	
Air Blank		0.000	19:45	
Diagnostics Check		OK	19:45	

Data Download Date/Time: 01/03/2020 05:19

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DEPUTY,.
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: FARINACCI,DEPUTY
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 19:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:49

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 12/31/2019 21:23:04

Diagnostics Check		OK	21:16
Air Blank		0.000	21:17
Control Test		0.083	21:17
Air Blank		0.000	21:18
Subject Sample	# 1	0.000	21:18
Breath Volume		1.453	21:18
Air Blank		0.000	21:19
Air Blank		0.000	21:21
Subject Sample	# 2	0.000	21:21
Breath Volume		1.296	21:21
Air Blank		0.000	21:22
Control Test		0.083	21:22
Air Blank		0.000	21:22
Diagnostics Check		OK	21:23

Data Download Date/Time: 01/03/2020 05:19

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASSEY,DARRELL,L
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: VREEKEN,RANDON
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 10:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:50

SUBJECT'S LAST NAME: BELTRAN

DATE/TIME OF TEST: 12/07/2019 11:20:10

Diagnostics Check		OK	11:17	
Air Blank		0.000	11:17	
Control Test		0.081	11:18	
Air Blank		0.000	11:18	
Subject Sample	# 1		11:18	Subject Test Refused
Breath Volume		0.000	11:18	
Air Blank		0.000	11:19	
Control Test		0.082	11:19	
Air Blank		0.000	11:20	
Diagnostics Check		OK	11:20	

Data Download Date/Time: 12/26/2019 12:27

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VEGA,JASON,
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: BOGUSKY,DENNIS
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: VARGAS-SANCHEZ

DATE/TIME OF TEST: 12/14/2019 03:49:26

Diagnostics Check		OK	03:43
Air Blank		0.000	03:43
Control Test		0.081	03:44
Air Blank		0.000	03:44
Subject Sample	# 1	0.173	03:44
Breath Volume		3.621	03:44
Air Blank		0.000	03:45
Air Blank		0.000	03:47
Subject Sample	# 2	0.167	03:47
Breath Volume		3.722	03:47
Air Blank		0.000	03:48
Control Test		0.080	03:48
Air Blank		0.000	03:49
Diagnostics Check		OK	03:49

Data Download Date/Time: 12/26/2019 12:27

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VEGA,JASON,
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: GARZA,DAVID
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: ONUCHOVSKY

DATE/TIME OF TEST: 12/14/2019 04:31:59

Diagnostics Check		OK	04:29	
Air Blank		0.000	04:29	
Control Test		0.081	04:29	
Air Blank		0.000	04:30	
Subject Sample	# 1		04:30	Subject Test Refused
Breath Volume		0.000	04:30	
Air Blank		0.000	04:31	
Control Test		0.081	04:31	
Air Blank		0.000	04:31	
Diagnostics Check		OK	04:31	

Data Download Date/Time: 12/26/2019 12:27

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,JOHN,P
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: GAMMAD,GARY
ARRESTING OFFICER AGENCY: Arcadia PD
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: RUIZ GARCIA

DATE/TIME OF TEST: 12/20/2019 20:58:05

Diagnostics Check		OK	20:49
Air Blank		0.000	20:50
Control Test		0.081	20:50
Air Blank		0.000	20:51
Subject Sample	# 1	0.245	20:53
Breath Volume		2.589	20:53
Air Blank		0.000	20:54
Air Blank		0.000	20:56
Subject Sample	# 2	0.226	20:56
Breath Volume		1.382	20:56
Air Blank		0.000	20:57
Control Test		0.080	20:57
Air Blank		0.000	20:58
Diagnostics Check		OK	20:58

Data Download Date/Time: 12/26/2019 12:27

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUEHLER,DEP,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BUEHLER,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: KECKLER

DATE/TIME OF TEST: 12/03/2019 22:51:17

Diagnostics Check		OK	22:45
Air Blank		0.000	22:45
Control Test		0.080	22:45
Air Blank		0.000	22:46
Subject Sample	# 1	0.284	22:46
Breath Volume		1.996	22:46
Air Blank		0.000	22:47
Air Blank		0.000	22:49
Subject Sample	# 2	0.280	22:49
Breath Volume		2.269	22:49
Air Blank		0.000	22:50
Control Test		0.079	22:50
Air Blank		0.000	22:51
Diagnostics Check		OK	22:51

Data Download Date/Time: 12/18/2019 10:56

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEDOREK,JEFFREY,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: CONLY,A
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: OLSEN

DATE/TIME OF TEST: 12/04/2019 04:18:00

Diagnostics Check		OK	04:11
Air Blank		0.000	04:12
Control Test		0.080	04:12
Air Blank		0.000	04:13
Subject Sample	# 1	0.221	04:13
Breath Volume		2.015	04:13
Air Blank		0.000	04:14
Air Blank		0.000	04:15
Subject Sample	# 2	0.218	04:16
Breath Volume		1.984	04:16
Air Blank		0.000	04:17
Control Test		0.079	04:17
Air Blank		0.000	04:17
Diagnostics Check		OK	04:18

Data Download Date/Time: 12/18/2019 10:56

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMBRIA,STEVEN,J
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: CAMBRIA,STEVEN
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 10:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:25

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 12/04/2019 12:57:51

Diagnostics Check		OK	12:51
Air Blank		0.000	12:51
Control Test		0.078	12:52
Air Blank		0.000	12:52
Subject Sample	# 1	0.000	12:53
Breath Volume		2.835	12:53
Air Blank		0.000	12:53
Air Blank		0.000	12:55
Subject Sample	# 2	0.000	12:56
Breath Volume		1.484	12:56
Air Blank		0.000	12:56
Control Test		0.079	12:57
Air Blank		0.000	12:57
Diagnostics Check		OK	12:57

Data Download Date/Time: 12/18/2019 10:56

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLYNN,MICHAEL,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BELL,D
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: COLON

DATE/TIME OF TEST: 12/04/2019 22:23:10

Diagnostics Check		OK	22:16
Air Blank		0.000	22:17
Control Test		0.080	22:17
Air Blank		0.000	22:17
Subject Sample	# 1	0.285	22:18
Breath Volume		1.957	22:18
Air Blank		0.000	22:19
Air Blank		0.000	22:21
Subject Sample	# 2	0.273	22:21
Breath Volume		1.570	22:21
Air Blank		0.000	22:22
Control Test		0.078	22:22
Air Blank		0.000	22:23
Diagnostics Check		OK	22:23

Data Download Date/Time: 12/18/2019 10:56

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERKINS,OFC,W
BREATH TEST OPERATOR AGENCY: Sarasota PD
ARRESTING OFFICER: PERKINS,OFC
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: POWELL

DATE/TIME OF TEST: 12/05/2019 00:47:10

Diagnostics Check		OK	00:40
Air Blank		0.000	00:41
Control Test		0.079	00:41
Air Blank		0.000	00:41
Subject Sample	# 1	0.000	00:42
Breath Volume		2.082	00:42
Air Blank		0.000	00:43
Air Blank		0.000	00:45
Subject Sample	# 2	0.000	00:45
Breath Volume		2.035	00:45
Air Blank		0.000	00:46
Control Test		0.079	00:46
Air Blank		0.000	00:47
Diagnostics Check		OK	00:47

Data Download Date/Time: 12/18/2019 10:56

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDEZ,DEPUTY,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BELL,DEPUTY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: AGOSTO

DATE/TIME OF TEST: 12/05/2019 03:25:56

Diagnostics Check		OK	03:19
Air Blank		0.000	03:19
Control Test		0.080	03:19
Air Blank		0.000	03:20
Subject Sample	# 1	0.229	03:21
Breath Volume		2.480	03:21
Air Blank		0.000	03:21
Air Blank		0.000	03:23
Subject Sample	# 2	0.225	03:24
Breath Volume		1.621	03:24
Air Blank		0.000	03:25
Control Test		0.079	03:25
Air Blank		0.000	03:25
Diagnostics Check		OK	03:25

Data Download Date/Time: 12/18/2019 10:56

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,STEVE,P
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SHORT,STEVEN
ARRESTING OFFICER AGENCY: Venice PD
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: HANCY

DATE/TIME OF TEST: 12/07/2019 05:30:31

Diagnostics Check		OK	05:22
Air Blank		0.000	05:23
Control Test		0.080	05:23
Air Blank		0.000	05:23
Subject Sample	# 1	0.137	05:24
Breath Volume		2.890	05:24
Air Blank		0.000	05:25
Air Blank		0.000	05:26
Subject Sample	# 2	0.136	05:28
Breath Volume		3.164	05:28
Air Blank		0.000	05:29
Control Test		0.079	05:29
Air Blank		0.000	05:30
Diagnostics Check		OK	05:30

Data Download Date/Time: 12/18/2019 10:56

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BOWERS,JOSEPH
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: BOYD

DATE/TIME OF TEST: 12/08/2019 01:26:49

Diagnostics Check		OK	01:20
Air Blank		0.000	01:20
Control Test		0.080	01:21
Air Blank		0.000	01:21
Subject Sample	# 1	0.179	01:22
Breath Volume		1.542	01:22
Air Blank		0.000	01:23
Air Blank		0.000	01:24
Subject Sample	# 2	0.184	01:25
Breath Volume		1.582	01:25
Air Blank		0.000	01:25
Control Test		0.078	01:26
Air Blank		0.000	01:26
Diagnostics Check		OK	01:26

Data Download Date/Time: 12/18/2019 10:56

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOETGER,ELI,L
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: BOETGER,ELI
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: CASIMIR

DATE/TIME OF TEST: 12/11/2019 03:33:00

Diagnostics Check		OK	03:25
Air Blank		0.000	03:26
Control Test		0.079	03:26
Air Blank		0.000	03:27
Subject Sample	# 1	0.117	03:28
Breath Volume		1.277	03:28
Air Blank		0.000	03:29
Air Blank		0.000	03:31
Subject Sample	# 2	0.115	03:31
Breath Volume		3.066	03:31
Air Blank		0.000	03:32
Control Test		0.078	03:32
Air Blank		0.000	03:32
Diagnostics Check		OK	03:33

Data Download Date/Time: 12/18/2019 10:56

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BOWERS,JOSEPH
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: HOLDA

DATE/TIME OF TEST: 12/12/2019 00:07:33

Diagnostics Check		OK	00:01
Air Blank		0.000	00:01
Control Test		0.079	00:02
Air Blank		0.000	00:02
Subject Sample	# 1	0.209	00:03
Breath Volume		1.550	00:03
Air Blank		0.000	00:03
Air Blank		0.000	00:05
Subject Sample	# 2	0.217	00:05
Breath Volume		2.304	00:05
Air Blank		0.000	00:06
Control Test		0.078	00:06
Air Blank		0.000	00:07
Diagnostics Check		OK	00:07

Data Download Date/Time: 12/18/2019 10:56

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACOSTA-BUGOS,KASSANDRA,J
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ACOSTA-BURGOS,KASSANDRA
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:15

SUBJECT'S LAST NAME: JOSEPH

DATE/TIME OF TEST: 12/12/2019 06:46:13

Diagnostics Check		OK	06:36	
Air Blank		0.000	06:36	
Control Test		0.080	06:37	
Air Blank		0.000	06:37	
Subject Sample	# 1		06:40	No Sample Provided
Breath Volume		0.000	06:40	
Air Blank		0.000	06:41	
Air Blank		0.000	06:43	
Subject Sample	# 2		06:44	Subject Test Refused
Breath Volume		0.000	06:44	
Air Blank		0.000	06:45	
Control Test		0.080	06:45	
Air Blank		0.000	06:46	
Diagnostics Check		OK	06:46	

Data Download Date/Time: 12/18/2019 10:56

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,TIMOTHY,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,TIMOTHY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: COUCH II

DATE/TIME OF TEST: 12/13/2019 21:26:40

Diagnostics Check		OK	21:20
Air Blank		0.000	21:20
Control Test		0.078	21:21
Air Blank		0.000	21:21
Subject Sample	# 1	0.151	21:22
Breath Volume		2.531	21:22
Air Blank		0.000	21:22
Air Blank		0.000	21:24
Subject Sample	# 2	0.142	21:25
Breath Volume		2.441	21:25
Air Blank		0.000	21:25
Control Test		0.079	21:26
Air Blank		0.000	21:26
Diagnostics Check		OK	21:26

Data Download Date/Time: 12/18/2019 10:56

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDEZ,DEPUTY,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BELL,DEPUTY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 12/14/2019 00:53:15

Diagnostics Check		OK	00:47
Air Blank		0.000	00:47
Control Test		0.079	00:47
Air Blank		0.000	00:48
Subject Sample	# 1	0.106	00:48
Breath Volume		3.343	00:48
Air Blank		0.000	00:49
Air Blank		0.000	00:51
Subject Sample	# 2	0.105	00:51
Breath Volume		3.125	00:51
Air Blank		0.000	00:52
Control Test		0.079	00:52
Air Blank		0.000	00:53
Diagnostics Check		OK	00:53

Data Download Date/Time: 12/18/2019 10:56

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUTLER,CHRIS,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BUTLER,CHRIS
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: VIGORELLI

DATE/TIME OF TEST: 12/14/2019 03:27:57

Diagnostics Check		OK	03:21
Air Blank		0.000	03:21
Control Test		0.078	03:21
Air Blank		0.000	03:22
Subject Sample	# 1	0.164	03:23
Breath Volume		1.128	03:23
Air Blank		0.000	03:23
Air Blank		0.000	03:25
Subject Sample	# 2	0.148	03:26
Breath Volume		1.316	03:26
Air Blank		0.000	03:27
Control Test		0.075	03:27
Air Blank		0.000	03:27
Diagnostics Check		OK	03:27

Data Download Date/Time: 12/18/2019 10:56

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDEZ,DEPUTY,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BELL,DEPUTY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: SILVA

DATE/TIME OF TEST: 12/14/2019 23:49:22

Diagnostics Check		OK	23:43
Air Blank		0.000	23:43
Control Test		0.078	23:43
Air Blank		0.000	23:44
Subject Sample	# 1	0.189	23:44
Breath Volume		3.519	23:44
Air Blank		0.000	23:45
Air Blank		0.000	23:47
Subject Sample	# 2	0.196	23:47
Breath Volume		2.429	23:47
Air Blank		0.000	23:48
Control Test		0.078	23:48
Air Blank		0.000	23:49
Diagnostics Check		OK	23:49

Data Download Date/Time: 12/18/2019 10:56

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDEZ,DEPUTY,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BELL,DEPUTY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: SILVA

DATE/TIME OF TEST: 12/14/2019 23:58:40

Diagnostics Check		OK	23:52
Air Blank		0.000	23:53
Control Test		0.079	23:53
Air Blank		0.000	23:53
Subject Sample	# 1	0.194	23:54
Breath Volume		2.554	23:54
Air Blank		0.000	23:54
Air Blank		0.000	23:56
Subject Sample	# 2	0.189	23:57
Breath Volume		2.121	23:57
Air Blank		0.000	23:57
Control Test		0.079	23:58
Air Blank		0.000	23:58
Diagnostics Check		OK	23:58

Data Download Date/Time: 12/18/2019 10:56

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,TIMOTHY,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,TIMOTHY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: WENTWORTH

DATE/TIME OF TEST: 12/15/2019 02:06:20

Diagnostics Check		OK	02:00
Air Blank		0.000	02:00
Control Test		0.080	02:01
Air Blank		0.000	02:01
Subject Sample	# 1	0.199	02:02
Breath Volume		2.042	02:02
Air Blank		0.000	02:02
Air Blank		0.000	02:04
Subject Sample	# 2	0.208	02:04
Breath Volume		1.722	02:04
Air Blank		0.000	02:05
Control Test		0.079	02:05
Air Blank		0.000	02:06
Diagnostics Check		OK	02:06

Data Download Date/Time: 12/18/2019 10:56

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANTWELL,J,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: CANTWELL,J
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 06:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:45

SUBJECT'S LAST NAME: PEDRO

DATE/TIME OF TEST: 12/15/2019 07:12:19

Diagnostics Check		OK	07:05
Air Blank		0.000	07:06
Control Test		0.080	07:06
Air Blank		0.000	07:07
Subject Sample	# 1	0.103	07:07
Breath Volume		1.585	07:07
Air Blank		0.000	07:08
Air Blank		0.000	07:10
Subject Sample	# 2	0.096	07:10
Breath Volume		2.449	07:10
Air Blank		0.000	07:11
Control Test		0.079	07:11
Air Blank		0.000	07:12
Diagnostics Check		OK	07:12

Data Download Date/Time: 12/18/2019 10:56

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAMORS,JASON,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SKINNER,D
ARRESTING OFFICER AGENCY: Venice PD
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: MCMAHON

DATE/TIME OF TEST: 12/16/2019 01:43:24

Diagnostics Check		OK	01:34	
Air Blank		0.000	01:34	
Control Test		0.080	01:35	
Air Blank		0.000	01:35	
Subject Sample	# 1		01:38	No Sample Provided
Breath Volume		0.000	01:38	
Air Blank		0.000	01:39	
Air Blank		0.000	01:41	
Subject Sample	# 2		01:42	Subject Test Refused
Breath Volume		0.000	01:42	
Air Blank		0.000	01:42	
Control Test		0.079	01:42	
Air Blank		0.000	01:43	
Diagnostics Check		OK	01:43	

Data Download Date/Time: 12/18/2019 10:56

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STROUP,STEVE,A
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: STARLING,WILLIAM
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 12/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: COFFIN

DATE/TIME OF TEST: 12/15/2019 21:59:54

Diagnostics Check		OK	21:53
Air Blank		0.000	21:53
Control Test		0.081	21:54
Air Blank		0.000	21:54
Subject Sample	# 1	0.212	21:55
Breath Volume		4.667	21:55
Air Blank		0.000	21:56
Air Blank		0.000	21:57
Subject Sample	# 2	0.216	21:58
Breath Volume		5.324	21:58
Air Blank		0.000	21:58
Control Test		0.079	21:59
Air Blank		0.000	21:59
Diagnostics Check		OK	21:59

Data Download Date/Time: 01/21/2020 04:55

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HACKETT,SEAN,J
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: KRONDHAL,CARL
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 12/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: HAMILTON

DATE/TIME OF TEST: 12/22/2019 23:56:16

Diagnostics Check		OK	23:48
Air Blank		0.000	23:49
Control Test		0.080	23:49
Air Blank		0.000	23:50
Subject Sample	# 1	0.169	23:51
Breath Volume		1.710	23:51
Air Blank		0.000	23:51
Air Blank		0.000	23:53
Subject Sample	# 2	0.173	23:54
Breath Volume		1.941	23:54
Air Blank		0.000	23:55
Control Test		0.078	23:55
Air Blank		0.000	23:56
Diagnostics Check		OK	23:56

Data Download Date/Time: 01/21/2020 04:55

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DEP,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MILLER,DEP R
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: VICCARO

DATE/TIME OF TEST: 12/04/2019 22:26:23

Diagnostics Check		OK	22:20	
Air Blank		0.000	22:20	
Control Test		0.080	22:21	
Air Blank		0.000	22:21	
Subject Sample	# 1	0.188	22:22	
Breath Volume		3.312	22:22	
Air Blank		0.000	22:22	
Air Blank		0.000	22:24	
Subject Sample	# 2	0.184	22:24	
Breath Volume		2.996	22:24	
Air Blank		0.000	22:25	
Control Test		0.087	22:25	Control Outside Tolerance
Air Blank		0.000	22:26	

Data Download Date/Time: 12/18/2019 11:45

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DEP,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MILLER,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: RUSSO

DATE/TIME OF TEST: 12/10/2019 03:34:37

Diagnostics Check		OK	03:28
Air Blank		0.000	03:28
Control Test		0.080	03:29
Air Blank		0.000	03:29
Subject Sample	# 1	0.182	03:30
Breath Volume		2.195	03:30
Air Blank		0.000	03:30
Air Blank		0.000	03:32
Subject Sample	# 2	0.184	03:33
Breath Volume		2.265	03:33
Air Blank		0.000	03:33
Control Test		0.079	03:34
Air Blank		0.000	03:34
Diagnostics Check		OK	03:34

Data Download Date/Time: 12/18/2019 11:45

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DEP,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BELL,DEP D
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:45

SUBJECT'S LAST NAME: EDGE

DATE/TIME OF TEST: 12/10/2019 17:27:44

Diagnostics Check		OK	17:21
Air Blank		0.000	17:21
Control Test		0.079	17:21
Air Blank		0.000	17:22
Subject Sample	# 1	0.198	17:23
Breath Volume		1.378	17:23
Air Blank		0.000	17:23
Air Blank		0.000	17:25
Subject Sample	# 2	0.194	17:26
Breath Volume		1.425	17:26
Air Blank		0.000	17:26
Control Test		0.079	17:27
Air Blank		0.000	17:27
Diagnostics Check		OK	17:27

Data Download Date/Time: 12/18/2019 11:45

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DEP,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MILLER,DEP R
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: NAVARRETE

DATE/TIME OF TEST: 12/15/2019 21:09:18

Diagnostics Check		OK	21:01
Air Blank		0.000	21:01
Control Test		0.080	21:02
Air Blank		0.000	21:02
Subject Sample	# 1	0.060	21:03
Breath Volume		1.492	21:03
Air Blank		0.000	21:03
Air Blank		0.000	21:05
Subject Sample	# 2	0.063	21:07
Breath Volume		1.203	21:07
Air Blank		0.000	21:08
Control Test		0.080	21:08
Air Blank		0.000	21:09
Diagnostics Check		OK	21:09

Data Download Date/Time: 12/18/2019 11:45

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DEP,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MILLER,DEP R
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: ZIMDARS

DATE/TIME OF TEST: 12/23/2019 23:35:36

Diagnostics Check		OK	23:27
Air Blank		0.000	23:28
Control Test		0.080	23:28
Air Blank		0.000	23:29
Subject Sample	# 1	0.277	23:30
Breath Volume		1.593	23:30
Air Blank		0.000	23:31
Air Blank		0.000	23:32
Subject Sample	# 2	0.260	23:34
Breath Volume		1.480	23:34
Air Blank		0.000	23:34
Control Test		0.078	23:35
Air Blank		0.000	23:35
Diagnostics Check		OK	23:35

Data Download Date/Time: 01/21/2020 09:55

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DEP,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MILLER,DEP R
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: CHISM

DATE/TIME OF TEST: 12/24/2019 22:54:53

Diagnostics Check		OK	22:48
Air Blank		0.000	22:49
Control Test		0.079	22:49
Air Blank		0.000	22:50
Subject Sample	# 1	0.271	22:50
Breath Volume		3.105	22:50
Air Blank		0.000	22:51
Air Blank		0.000	22:52
Subject Sample	# 2	0.261	22:53
Breath Volume		1.523	22:53
Air Blank		0.000	22:53
Control Test		0.079	22:54
Air Blank		0.000	22:54
Diagnostics Check		OK	22:54

Data Download Date/Time: 01/21/2020 09:55

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DEP,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MILLER,DEP R
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: CHISM

DATE/TIME OF TEST: 12/24/2019 23:44:25

Diagnostics Check		OK	23:38
Air Blank		0.000	23:38
Control Test		0.079	23:39
Air Blank		0.000	23:39
Subject Sample	# 1	0.255	23:40
Breath Volume		2.929	23:40
Air Blank		0.000	23:40
Air Blank		0.000	23:42
Subject Sample	# 2	0.257	23:42
Breath Volume		3.242	23:42
Air Blank		0.000	23:43
Control Test		0.077	23:43
Air Blank		0.000	23:44
Diagnostics Check		OK	23:44

Data Download Date/Time: 01/21/2020 09:55

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEPUTY,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEPUTY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:46

SUBJECT'S LAST NAME: PALOMAA

DATE/TIME OF TEST: 12/18/2019 19:12:52

Diagnostics Check		OK	19:06
Air Blank		0.000	19:07
Control Test		0.079	19:07
Air Blank		0.000	19:08
Subject Sample	# 1	0.165	19:08
Breath Volume		2.222	19:08
Air Blank		0.000	19:09
Air Blank		0.000	19:11
Subject Sample	# 2	0.148	19:11
Breath Volume		1.269	19:11
Air Blank		0.000	19:11
Control Test		0.079	19:12
Air Blank		0.000	19:12
Diagnostics Check		OK	19:12

Data Download Date/Time: 01/21/2020 08:05

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEPUTY,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEPUTY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: MATTHEWS

DATE/TIME OF TEST: 12/19/2019 03:31:54

Diagnostics Check		OK	03:20	
Air Blank		0.000	03:21	
Control Test		0.080	03:21	
Air Blank		0.000	03:21	
Subject Sample	# 1		03:24	No Sample Provided
Breath Volume		0.000	03:24	
Air Blank		0.000	03:25	
Air Blank		0.000	03:27	
Subject Sample	# 2		03:30	No Sample Provided
Breath Volume		0.000	03:30	
Air Blank		0.000	03:31	
Control Test		0.081	03:31	
Air Blank		0.000	03:31	
Diagnostics Check		OK	03:31	

Data Download Date/Time: 01/21/2020 08:05

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DARLING,JOHN,P
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: RICHARDS,J
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 12/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 06:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:00

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 12/19/2019 08:43:24

Diagnostics Check		OK	08:37
Air Blank		0.000	08:37
Control Test		0.081	08:38
Air Blank		0.000	08:38
Subject Sample	# 1	0.171	08:39
Breath Volume		2.187	08:39
Air Blank		0.000	08:39
Air Blank		0.000	08:41
Subject Sample	# 2	0.173	08:41
Breath Volume		2.527	08:41
Air Blank		0.000	08:42
Control Test		0.079	08:42
Air Blank		0.000	08:43
Diagnostics Check		OK	08:43

Data Download Date/Time: 01/21/2020 08:05

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,J,S
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BARR,J
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: GRILL

DATE/TIME OF TEST: 12/20/2019 22:00:20

Diagnostics Check		OK	21:51	
Air Blank		0.000	21:51	
Control Test		0.079	21:51	
Air Blank		0.000	21:52	
Subject Sample	# 1		21:55	No Sample Provided
Breath Volume		0.000	21:55	
Air Blank		0.000	21:55	
Air Blank		0.000	21:57	
Subject Sample	# 2		21:59	RFI Detect
Breath Volume		0.000	21:59	
Air Blank			21:59	RFI Detect
Air Blank		0.000	22:00	

Data Download Date/Time: 01/21/2020 08:05

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FELTOVIC,DEP,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: GRIFFIN,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 13:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:04

SUBJECT'S LAST NAME: LOVELY

DATE/TIME OF TEST: 12/21/2019 15:33:37

Diagnostics Check		OK	15:27
Air Blank		0.000	15:28
Control Test		0.079	15:28
Air Blank		0.000	15:29
Subject Sample	# 1	0.252	15:29
Breath Volume		3.843	15:29
Air Blank		0.000	15:29
Air Blank		0.000	15:31
Subject Sample	# 2	0.253	15:32
Breath Volume		2.410	15:32
Air Blank		0.000	15:32
Control Test		0.077	15:33
Air Blank		0.000	15:33
Diagnostics Check		OK	15:33

Data Download Date/Time: 01/21/2020 08:05

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,STEVE,P
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: ESMOND,A
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 12/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: VILLA SANCHEZ

DATE/TIME OF TEST: 12/23/2019 05:13:15

Diagnostics Check		OK	05:07
Air Blank		0.000	05:07
Control Test		0.081	05:08
Air Blank		0.000	05:08
Subject Sample	# 1	0.127	05:08
Breath Volume		2.144	05:08
Air Blank		0.000	05:09
Air Blank		0.000	05:11
Subject Sample	# 2	0.121	05:11
Breath Volume		1.554	05:11
Air Blank		0.000	05:12
Control Test		0.079	05:12
Air Blank		0.000	05:13
Diagnostics Check		OK	05:13

Data Download Date/Time: 01/21/2020 08:05

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDEZ,DEPUTY,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BELL,DEPUTY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: JOHNSTON

DATE/TIME OF TEST: 12/23/2019 23:31:34

Diagnostics Check		OK	23:25
Air Blank		0.000	23:25
Control Test		0.080	23:26
Air Blank		0.000	23:26
Subject Sample	# 1	0.214	23:27
Breath Volume		2.074	23:27
Air Blank		0.000	23:27
Air Blank		0.000	23:29
Subject Sample	# 2	0.215	23:30
Breath Volume		1.625	23:30
Air Blank		0.000	23:30
Control Test		0.080	23:30
Air Blank		0.000	23:31
Diagnostics Check		OK	23:31

Data Download Date/Time: 01/21/2020 08:05

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACOSTA-BURGOS,KASSANDRA,J
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ACOSTA-BURGOS,KASSANDRA
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male

OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: GIFFARD

DATE/TIME OF TEST: 12/25/2019 22:50:54

Diagnostics Check		OK	22:44
Air Blank		0.000	22:44
Control Test		0.080	22:45
Air Blank		0.000	22:45
Subject Sample	# 1	0.186	22:46
Breath Volume		2.968	22:46
Air Blank		0.000	22:47
Air Blank		0.000	22:48
Subject Sample	# 2	0.172	22:49
Breath Volume		2.722	22:49
Air Blank		0.000	22:50
Control Test		0.079	22:50
Air Blank		0.000	22:50
Diagnostics Check		OK	22:50

Data Download Date/Time: 01/21/2020 08:05

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEPUTY,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,T
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: DOYLE

DATE/TIME OF TEST: 12/27/2019 22:46:53

Diagnostics Check		OK	22:40
Air Blank		0.000	22:41
Control Test		0.080	22:41
Air Blank		0.000	22:42
Subject Sample	# 1	0.196	22:42
Breath Volume		1.257	22:42
Air Blank		0.000	22:43
Air Blank		0.000	22:45
Subject Sample	# 2	0.196	22:45
Breath Volume		2.253	22:45
Air Blank		0.000	22:45
Control Test		0.080	22:46
Air Blank		0.000	22:46
Diagnostics Check		OK	22:46

Data Download Date/Time: 01/21/2020 08:05

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDEZ,DEPUTY,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: ST.JEAN,OFFICER
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 12/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: GUSTAFSON

DATE/TIME OF TEST: 12/28/2019 03:29:31

Diagnostics Check		OK	03:22
Air Blank		0.000	03:23
Control Test		0.080	03:23
Air Blank		0.000	03:24
Subject Sample	# 1	0.101	03:24
Breath Volume		3.488	03:24
Air Blank		0.000	03:25
Air Blank		0.000	03:27
Subject Sample	# 2	0.093	03:28
Breath Volume		1.113	03:28
Air Blank		0.000	03:28
Control Test		0.080	03:28
Air Blank		0.000	03:29
Diagnostics Check		OK	03:29

Data Download Date/Time: 01/21/2020 08:05

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDEZ,DEPUTY,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BELL,DEPUTY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: SIRUS

DATE/TIME OF TEST: 12/28/2019 04:38:28

Diagnostics Check		OK	04:32
Air Blank		0.000	04:32
Control Test		0.078	04:33
Air Blank		0.000	04:33
Subject Sample	# 1	0.220	04:34
Breath Volume		2.148	04:34
Air Blank		0.000	04:34
Air Blank		0.000	04:36
Subject Sample	# 2	0.219	04:37
Breath Volume		1.574	04:37
Air Blank		0.000	04:37
Control Test		0.079	04:37
Air Blank		0.000	04:38
Diagnostics Check		OK	04:38

Data Download Date/Time: 01/21/2020 08:05

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEPUTY,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEPUTY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 12/30/2019 03:46:27

Diagnostics Check		OK	03:40
Air Blank		0.000	03:41
Control Test		0.079	03:41
Air Blank		0.000	03:41
Subject Sample	# 1	0.190	03:42
Breath Volume		1.218	03:42
Air Blank		0.000	03:42
Air Blank		0.000	03:44
Subject Sample	# 2	0.196	03:45
Breath Volume		1.199	03:45
Air Blank		0.000	03:45
Control Test		0.079	03:45
Air Blank		0.000	03:46
Diagnostics Check		OK	03:46

Data Download Date/Time: 01/21/2020 08:05

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,STEVEN,P
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: COKER,JOSH
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 12/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: SAVCHENKO

DATE/TIME OF TEST: 12/31/2019 01:08:24

Diagnostics Check	OK	01:06	
Air Blank	0.000	01:07	
Control Test	0.080	01:07	
Air Blank		01:07	RFI Detect
Air Blank	0.000	01:08	

Data Download Date/Time: 01/21/2020 08:05

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,STEVEN,P
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: COKER,JOSH
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 12/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: SAVCHENKO

DATE/TIME OF TEST: 12/31/2019 01:26:10

Diagnostics Check		OK	01:10	
Air Blank		0.000	01:10	
Control Test		0.080	01:11	
Air Blank		0.000	01:11	
Subject Sample	# 1	0.141	01:14	Volume Not Met
Breath Volume		0.445	01:14	
Air Blank		0.000	01:15	
Air Blank		0.000	01:17	
Subject Sample	# 2	0.178	01:19	
Breath Volume		2.437	01:19	
Air Blank		0.000	01:19	
Air Blank		0.000	01:21	
Subject Sample	# 3		01:24	No Sample Provided
Breath Volume		0.000	01:24	
Air Blank		0.000	01:25	
Control Test		0.076	01:25	
Air Blank		0.000	01:26	
Diagnostics Check		OK	01:26	

Data Download Date/Time: 01/21/2020 08:05

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,J,S
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BARR,J
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: POWERS

DATE/TIME OF TEST: 12/02/2019 23:51:13

Diagnostics Check		OK	23:44
Air Blank		0.000	23:44
Control Test		0.080	23:45
Air Blank		0.000	23:45
Subject Sample	# 1	0.217	23:46
Breath Volume		3.695	23:46
Air Blank		0.000	23:47
Air Blank		0.000	23:49
Subject Sample	# 2	0.213	23:49
Breath Volume		1.941	23:49
Air Blank		0.000	23:50
Control Test		0.079	23:50
Air Blank		0.000	23:51
Diagnostics Check		OK	23:51

Data Download Date/Time: 12/18/2019 11:00

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ECKERT,GABRIEL,B
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: ECKERT,GABRIEL
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: WILCOX

DATE/TIME OF TEST: 12/06/2019 21:32:37

Diagnostics Check		OK	21:26
Air Blank		0.000	21:26
Control Test		0.079	21:27
Air Blank		0.000	21:27
Subject Sample	# 1	0.142	21:28
Breath Volume		1.921	21:28
Air Blank		0.000	21:28
Air Blank		0.000	21:30
Subject Sample	# 2	0.147	21:31
Breath Volume		1.171	21:31
Air Blank		0.000	21:31
Control Test		0.079	21:32
Air Blank		0.000	21:32
Diagnostics Check		OK	21:32

Data Download Date/Time: 12/18/2019 11:00

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,J,S
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BARR,J
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: GREENLAW

DATE/TIME OF TEST: 12/06/2019 23:42:56

Diagnostics Check		OK	23:36
Air Blank		0.000	23:37
Control Test		0.079	23:37
Air Blank		0.000	23:38
Subject Sample	# 1	0.146	23:38
Breath Volume		1.710	23:38
Air Blank		0.000	23:39
Air Blank		0.000	23:40
Subject Sample	# 2	0.151	23:41
Breath Volume		1.582	23:41
Air Blank		0.000	23:42
Control Test		0.079	23:42
Air Blank		0.000	23:42
Diagnostics Check		OK	23:42

Data Download Date/Time: 12/18/2019 11:00

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACOSTA-BURGOS,KASSANDRA,J
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ACOSTA-BURGOS,KASSANDRA
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:59

SUBJECT'S LAST NAME: PEPLAU

DATE/TIME OF TEST: 12/12/2019 22:21:57

Diagnostics Check	OK	22:21	
Air Blank		22:21	RFI Detect
Air Blank	0.000	22:21	

Data Download Date/Time: 12/18/2019 11:00

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACOSTA-BURGOS,KASSANDRA,J
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ACOSTA-BURGOS,KASSANDRA
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Male

OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: PEPLAU

DATE/TIME OF TEST: 12/12/2019 22:32:43

Diagnostics Check		OK	22:25
Air Blank		0.000	22:26
Control Test		0.080	22:26
Air Blank		0.000	22:27
Subject Sample	# 1	0.066	22:28
Breath Volume		3.808	22:28
Air Blank		0.000	22:28
Air Blank		0.000	22:30
Subject Sample	# 2	0.065	22:31
Breath Volume		2.753	22:31
Air Blank		0.000	22:31
Control Test		0.079	22:32
Air Blank		0.000	22:32
Diagnostics Check		OK	22:32

Data Download Date/Time: 12/18/2019 11:00

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,J,S
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BARR,J
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: STASIK

DATE/TIME OF TEST: 12/22/2019 02:47:14

Diagnostics Check		OK	02:40
Air Blank		0.000	02:41
Control Test		0.080	02:41
Air Blank		0.000	02:42
Subject Sample	# 1	0.168	02:42
Breath Volume		2.882	02:42
Air Blank		0.000	02:43
Air Blank		0.000	02:45
Subject Sample	# 2	0.156	02:45
Breath Volume		3.089	02:45
Air Blank		0.000	02:46
Control Test		0.079	02:46
Air Blank		0.000	02:47
Diagnostics Check		OK	02:47

Data Download Date/Time: 01/21/2020 09:42

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,J,S
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: ACOSTA-BURGOS,KASSANDRA
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Female

OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: BONNER

DATE/TIME OF TEST: 12/23/2019 00:37:47

Diagnostics Check	OK	00:37	
Air Blank		00:37	RFI Detect
Air Blank	0.000	00:37	

Data Download Date/Time: 01/21/2020 09:42

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,J,S
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: ACOSTA-BURGOS,KASSANDRA
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Female

OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: BONNER

DATE/TIME OF TEST: 12/23/2019 00:45:26

Diagnostics Check		OK	00:39
Air Blank		0.000	00:39
Control Test		0.078	00:39
Air Blank		0.000	00:40
Subject Sample	# 1	0.197	00:40
Breath Volume		2.257	00:40
Air Blank		0.000	00:41
Air Blank		0.000	00:43
Subject Sample	# 2	0.186	00:43
Breath Volume		2.253	00:43
Air Blank		0.000	00:44
Control Test		0.079	00:44
Air Blank		0.000	00:45
Diagnostics Check		OK	00:45

Data Download Date/Time: 01/21/2020 09:42

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,J,S
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FLINT,C
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/18/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: ABARCA

DATE/TIME OF TEST: 12/26/2019 00:52:07

Diagnostics Check		OK	00:45
Air Blank		0.000	00:46
Control Test		0.079	00:46
Air Blank		0.000	00:47
Subject Sample	# 1	0.070	00:47
Breath Volume		4.714	00:47
Air Blank		0.000	00:48
Air Blank		0.000	00:50
Subject Sample	# 2	0.069	00:50
Breath Volume		4.792	00:50
Air Blank		0.000	00:51
Control Test		0.079	00:51
Air Blank		0.000	00:52
Diagnostics Check		OK	00:52

Data Download Date/Time: 01/21/2020 09:42

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAIR,TIMOTHY,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HITCHCOCK,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: CARROLL

DATE/TIME OF TEST: 12/02/2019 01:18:03

Diagnostics Check		OK	01:11
Air Blank		0.000	01:12
Control Test		0.081	01:12
Air Blank		0.000	01:12
Subject Sample	# 1	0.000	01:13
Breath Volume		2.121	01:13
Air Blank		0.000	01:14
Air Blank		0.000	01:16
Subject Sample	# 2	0.000	01:16
Breath Volume		1.621	01:16
Air Blank		0.000	01:17
Control Test		0.081	01:17
Air Blank		0.000	01:17
Diagnostics Check		OK	01:18

Data Download Date/Time: 01/02/2020 10:33

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK,WILLIAM,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HITCHCOCK,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 12/07/2019 05:31:20

Diagnostics Check		OK	05:24
Air Blank		0.000	05:25
Control Test		0.081	05:25
Air Blank		0.000	05:25
Subject Sample	# 1	0.184	05:27
Breath Volume		1.687	05:27
Air Blank		0.000	05:27
Air Blank		0.000	05:29
Subject Sample	# 2	0.182	05:29
Breath Volume		1.437	05:29
Air Blank		0.000	05:30
Control Test		0.080	05:30
Air Blank		0.000	05:31
Diagnostics Check		OK	05:31

Data Download Date/Time: 01/02/2020 10:33

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDEZ,JAMIE,J
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: SWISHER,EDWARD
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: HEE

DATE/TIME OF TEST: 12/14/2019 23:05:18

Diagnostics Check		OK	23:02	
Air Blank		0.000	23:02	
Control Test		0.081	23:03	
Air Blank		0.000	23:03	
Subject Sample	# 1		23:03	Subject Test Refused
Breath Volume		0.000	23:03	
Air Blank		0.000	23:04	
Control Test		0.081	23:04	
Air Blank		0.000	23:05	
Diagnostics Check		OK	23:05	

Data Download Date/Time: 01/02/2020 10:33

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWISHER,EDWARD,T
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: SWISHER,EDWARD
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 09:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:39

SUBJECT'S LAST NAME: SCHAD

DATE/TIME OF TEST: 12/28/2019 10:18:08

Diagnostics Check		OK	10:11
Air Blank		0.000	10:12
Control Test		0.080	10:12
Air Blank		0.000	10:13
Subject Sample	# 1	0.154	10:13
Breath Volume		4.238	10:13
Air Blank		0.000	10:14
Air Blank		0.000	10:16
Subject Sample	# 2	0.149	10:16
Breath Volume		3.906	10:16
Air Blank		0.000	10:17
Control Test		0.080	10:17
Air Blank		0.000	10:18
Diagnostics Check		OK	10:18

Data Download Date/Time: 01/02/2020 10:33

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZIGLER,RICKIE,R
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: ZIGLER,RICKIE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 12:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:55

SUBJECT'S LAST NAME: CYPHERS

DATE/TIME OF TEST: 12/31/2019 13:19:05

Diagnostics Check	OK	13:18	
Air Blank		13:18	RFI Detect
Air Blank		13:19	RFI Detect

Data Download Date/Time: 01/02/2020 10:33

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZIGLER,RICKIE,R
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: ZIGLER,RICKIE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 12:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:55

SUBJECT'S LAST NAME: CYPHERS

DATE/TIME OF TEST: 12/31/2019 13:28:30

Diagnostics Check		OK	13:21
Air Blank		0.000	13:22
Control Test		0.081	13:22
Air Blank		0.000	13:23
Subject Sample	# 1	0.328	13:23
Breath Volume		2.371	13:23
Air Blank		0.000	13:24
Air Blank		0.000	13:26
Subject Sample	# 2	0.322	13:26
Breath Volume		2.132	13:26
Air Blank		0.000	13:27
Control Test		0.079	13:27
Air Blank		0.000	13:28
Diagnostics Check		OK	13:28

Data Download Date/Time: 01/02/2020 10:33

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUYTON,JOHN,A
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GUYTON,JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 11/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: CHEVES

DATE/TIME OF TEST: 12/01/2019 02:37:37

Diagnostics Check		OK	02:31
Air Blank		0.000	02:32
Control Test		0.079	02:32
Air Blank		0.000	02:32
Subject Sample	# 1	0.098	02:33
Breath Volume		2.667	02:33
Air Blank		0.000	02:33
Air Blank		0.000	02:35
Subject Sample	# 2	0.092	02:36
Breath Volume		2.085	02:36
Air Blank		0.000	02:36
Control Test		0.079	02:37
Air Blank		0.000	02:37
Diagnostics Check		OK	02:37

Data Download Date/Time: 12/22/2019 04:49

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELCH,CHAD,J
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: SHRIGLEY,ROBERT
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 11/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: BLOWERS

DATE/TIME OF TEST: 12/01/2019 04:51:12

Diagnostics Check		OK	04:44
Air Blank		0.000	04:45
Control Test		0.078	04:45
Air Blank		0.000	04:46
Subject Sample	# 1	0.106	04:46
Breath Volume		1.742	04:46
Air Blank		0.000	04:47
Air Blank		0.000	04:49
Subject Sample	# 2	0.104	04:49
Breath Volume		1.636	04:49
Air Blank		0.000	04:50
Control Test		0.078	04:50
Air Blank		0.000	04:51
Diagnostics Check		OK	04:51

Data Download Date/Time: 12/22/2019 04:49

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUYTON,JOHN,A
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GUYTON,JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 11/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: ROLLINS

DATE/TIME OF TEST: 12/15/2019 03:27:36

Diagnostics Check		OK	03:21
Air Blank		0.000	03:21
Control Test		0.080	03:21
Air Blank		0.000	03:22
Subject Sample	# 1	0.097	03:23
Breath Volume		2.164	03:23
Air Blank		0.000	03:23
Air Blank		0.000	03:25
Subject Sample	# 2	0.093	03:26
Breath Volume		1.316	03:26
Air Blank		0.000	03:26
Control Test		0.080	03:27
Air Blank		0.000	03:27
Diagnostics Check		OK	03:27

Data Download Date/Time: 12/22/2019 04:49

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEIMIG,EMORIE,I
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: DOUALEHI,AHIPO
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 11/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: VELEZ VILLANUEVA

DATE/TIME OF TEST: 12/17/2019 04:12:25

Diagnostics Check		OK	04:05
Air Blank		0.000	04:06
Control Test		0.080	04:06
Air Blank		0.000	04:07
Subject Sample	# 1	0.211	04:07
Breath Volume		3.851	04:07
Air Blank		0.000	04:08
Air Blank		0.000	04:10
Subject Sample	# 2	0.208	04:10
Breath Volume		2.074	04:10
Air Blank		0.000	04:11
Control Test		0.079	04:11
Air Blank		0.000	04:12
Diagnostics Check		OK	04:12

Data Download Date/Time: 12/22/2019 04:49

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURWELL,MARK,J
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: BIVENS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 12/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: ASKERA

DATE/TIME OF TEST: 12/31/2019 02:19:35

Diagnostics Check	OK	02:17	
Air Blank	0.000	02:18	
Control Test	0.080	02:18	
Air Blank		02:19	RFI Detect
Air Blank	0.000	02:19	

Data Download Date/Time: 01/29/2020 01:42

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURWELL,MARK,J
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: BIVENS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 12/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: ASKERA

DATE/TIME OF TEST: 12/31/2019 02:28:13

Diagnostics Check		OK	02:21
Air Blank		0.000	02:22
Control Test		0.079	02:22
Air Blank		0.000	02:23
Subject Sample	# 1	0.177	02:23
Breath Volume		2.656	02:23
Air Blank		0.000	02:24
Air Blank		0.000	02:26
Subject Sample	# 2	0.166	02:26
Breath Volume		2.515	02:26
Air Blank		0.000	02:27
Control Test		0.080	02:27
Air Blank		0.000	02:28
Diagnostics Check		OK	02:28

Data Download Date/Time: 01/29/2020 01:42

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,SCOTT,C
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: DAVIS,KEVIN
ARRESTING OFFICER AGENCY: Levy County SO
LAST AGENCY INSPECTION: 11/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805701
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: LAVERDURE

DATE/TIME OF TEST: 12/05/2019 03:00:34

Diagnostics Check		OK	02:53
Air Blank		0.000	02:54
Control Test		0.079	02:54
Air Blank		0.000	02:54
Subject Sample	# 1	0.178	02:55
Breath Volume		2.531	02:55
Air Blank		0.000	02:56
Air Blank		0.000	02:58
Subject Sample	# 2	0.178	02:58
Breath Volume		2.144	02:58
Air Blank		0.000	02:59
Control Test		0.076	02:59
Air Blank		0.000	03:00
Diagnostics Check		OK	03:00

Data Download Date/Time: 12/16/2019 11:14

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,RICK,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BROWN,RICK
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 11/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805701
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 12:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: CRISS

DATE/TIME OF TEST: 12/14/2019 01:27:23

Diagnostics Check		OK	01:21
Air Blank		0.000	01:21
Control Test		0.079	01:22
Air Blank		0.000	01:22
Subject Sample	# 1	0.108	01:22
Breath Volume		1.515	01:22
Air Blank		0.000	01:23
Air Blank		0.000	01:25
Subject Sample	# 2	0.107	01:25
Breath Volume		2.031	01:25
Air Blank		0.000	01:26
Control Test		0.079	01:26
Air Blank		0.000	01:27
Diagnostics Check		OK	01:27

Data Download Date/Time: 12/16/2019 11:15

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MITCHELL,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 18:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:39

SUBJECT'S LAST NAME: AHLQUIST

DATE/TIME OF TEST: 12/18/2019 20:08:22

Diagnostics Check		OK	20:01
Air Blank		0.000	20:01
Control Test		0.080	20:02
Air Blank		0.000	20:02
Subject Sample	# 1	0.191	20:03
Breath Volume		1.851	20:03
Air Blank		0.000	20:04
Air Blank		0.000	20:05
Subject Sample	# 2	0.200	20:06
Breath Volume		1.593	20:06
Air Blank		0.000	20:07
Control Test		0.080	20:07
Air Blank		0.000	20:08
Diagnostics Check		OK	20:08

Data Download Date/Time: 12/21/2019 20:41

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: STASNEY

DATE/TIME OF TEST: 12/18/2019 22:59:06

Diagnostics Check		OK	22:56	
Air Blank		0.000	22:56	
Control Test		0.081	22:57	
Air Blank		0.000	22:57	
Subject Sample	# 1		22:57	Subject Test Refused
Breath Volume		0.000	22:57	
Air Blank		0.000	22:58	
Control Test		0.081	22:58	
Air Blank		0.000	22:59	
Diagnostics Check		OK	22:59	

Data Download Date/Time: 12/21/2019 20:41

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERGER,CHARLES,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BERGER,CHARLES
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 05:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:26

SUBJECT'S LAST NAME: MCCREARY

DATE/TIME OF TEST: 12/20/2019 05:49:22

Diagnostics Check		OK	05:46	
Air Blank		0.000	05:46	
Control Test		0.080	05:47	
Air Blank		0.000	05:47	
Subject Sample	# 1		05:47	Subject Test Refused
Breath Volume		0.000	05:47	
Air Blank		0.000	05:48	
Control Test		0.080	05:48	
Air Blank		0.000	05:49	
Diagnostics Check		OK	05:49	

Data Download Date/Time: 12/21/2019 20:41

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: GLOMB

DATE/TIME OF TEST: 12/20/2019 22:57:23

Diagnostics Check		OK	22:51
Air Blank		0.000	22:51
Control Test		0.081	22:52
Air Blank		0.000	22:52
Subject Sample	# 1	0.000	22:53
Breath Volume		2.441	22:53
Air Blank		0.000	22:53
Air Blank		0.000	22:55
Subject Sample	# 2	0.000	22:56
Breath Volume		2.152	22:56
Air Blank		0.000	22:56
Control Test		0.080	22:56
Air Blank		0.000	22:57
Diagnostics Check		OK	22:57

Data Download Date/Time: 12/21/2019 20:41

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,KYLE,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GRIFFIN,KYLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: WALSH

DATE/TIME OF TEST: 12/21/2019 02:39:14

Diagnostics Check		OK	02:27	
Air Blank		0.000	02:28	
Control Test		0.081	02:28	
Air Blank		0.000	02:29	
Subject Sample	# 1		02:32	No Sample Provided
Breath Volume		0.000	02:32	
Air Blank		0.000	02:32	
Air Blank		0.000	02:34	
Subject Sample	# 2		02:37	No Sample Provided
Breath Volume		0.000	02:37	
Air Blank		0.000	02:38	
Control Test		0.081	02:38	
Air Blank		0.000	02:39	
Diagnostics Check		OK	02:39	

Data Download Date/Time: 12/21/2019 20:41

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICOTTILLI,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI,JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: KENNY

DATE/TIME OF TEST: 12/21/2019 04:52:59

Diagnostics Check		OK	04:46
Air Blank		0.000	04:46
Control Test		0.081	04:47
Air Blank		0.000	04:47
Subject Sample	# 1	0.000	04:48
Breath Volume		3.230	04:48
Air Blank		0.000	04:48
Air Blank		0.000	04:50
Subject Sample	# 2	0.000	04:51
Breath Volume		3.632	04:51
Air Blank		0.000	04:52
Control Test		0.081	04:52
Air Blank		0.000	04:52
Diagnostics Check		OK	04:52

Data Download Date/Time: 12/21/2019 20:41

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANWORP,JUDSON,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: VANWORP,JUDSON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 04:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:08

SUBJECT'S LAST NAME: SWEET

DATE/TIME OF TEST: 12/21/2019 05:38:58

Diagnostics Check		OK	05:32
Air Blank		0.000	05:32
Control Test		0.081	05:32
Air Blank		0.000	05:33
Subject Sample	# 1	0.144	05:34
Breath Volume		1.203	05:34
Air Blank		0.000	05:34
Air Blank		0.000	05:36
Subject Sample	# 2	0.148	05:37
Breath Volume		3.527	05:37
Air Blank		0.000	05:38
Control Test		0.081	05:38
Air Blank		0.000	05:38
Diagnostics Check		OK	05:38

Data Download Date/Time: 12/21/2019 20:41

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SIMMONS,L
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 12/21/2019 21:51:53

Diagnostics Check		OK	21:49	
Air Blank		0.000	21:49	
Control Test		0.081	21:49	
Air Blank		0.000	21:50	
Subject Sample	# 1		21:50	Subject Test Refused
Breath Volume		0.000	21:50	
Air Blank		0.000	21:50	
Control Test		0.081	21:51	
Air Blank		0.000	21:51	
Diagnostics Check		OK	21:51	

Data Download Date/Time: 01/11/2020 21:07

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BOYLE,G
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 12/22/2019 05:01:07

Diagnostics Check		OK	04:55
Air Blank		0.000	04:55
Control Test		0.080	04:55
Air Blank		0.000	04:56
Subject Sample	# 1	0.000	04:56
Breath Volume		3.800	04:56
Air Blank		0.000	04:57
Air Blank		0.000	04:59
Subject Sample	# 2	0.000	04:59
Breath Volume		2.218	04:59
Air Blank		0.000	05:00
Control Test		0.080	05:00
Air Blank		0.000	05:01
Diagnostics Check		OK	05:01

Data Download Date/Time: 01/11/2020 21:07

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,ARTHUR
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: MCALEENAN

DATE/TIME OF TEST: 12/28/2019 03:11:27

Diagnostics Check		OK	03:08	
Air Blank		0.000	03:09	
Control Test		0.081	03:09	
Air Blank		0.000	03:09	
Subject Sample	# 1		03:10	Subject Test Refused
Breath Volume		0.000	03:10	
Air Blank		0.000	03:10	
Control Test		0.080	03:10	
Air Blank		0.000	03:11	
Diagnostics Check		OK	03:11	

Data Download Date/Time: 01/11/2020 21:07

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERGER,CHARLES,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BERGER,CHARLES
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 12/31/2019 23:50:52

Diagnostics Check		OK	23:37	
Air Blank		0.000	23:37	
Control Test		0.081	23:38	
Air Blank		0.000	23:38	
Subject Sample	# 1	0.166	23:40	No .020 agreement
Breath Volume		1.789	23:40	
Air Blank		0.000	23:40	
Air Blank		0.000	23:42	
Subject Sample	# 2	0.175	23:45	Volume Not Met
Breath Volume		0.218	23:45	
Air Blank		0.000	23:46	
Air Blank		0.000	23:48	
Subject Sample	# 3	0.196	23:49	No .020 agreement
Breath Volume		1.496	23:49	
Air Blank		0.000	23:49	
Control Test		0.078	23:50	
Air Blank		0.000	23:50	
Diagnostics Check		OK	23:50	

Data Download Date/Time: 01/11/2020 21:07

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001646

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: ADKINS,ANDREW
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 11/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 20:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 12/15/2019 21:01:55

Diagnostics Check		OK	20:55
Air Blank		0.000	20:56
Control Test		0.082	20:56
Air Blank		0.000	20:56
Subject Sample	# 1	0.107	20:57
Breath Volume		2.343	20:57
Air Blank		0.000	20:58
Air Blank		0.000	20:59
Subject Sample	# 2	0.106	21:00
Breath Volume		1.468	21:00
Air Blank		0.000	21:01
Control Test		0.082	21:01
Air Blank		0.000	21:01
Diagnostics Check		OK	21:01

Data Download Date/Time: 12/20/2019 12:35

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001646

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMBY,TAYLOR,N
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: VICK,ANTHONY
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 12/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: JIMENEZ

DATE/TIME OF TEST: 12/25/2019 02:59:42

Diagnostics Check		OK	02:53
Air Blank		0.000	02:53
Control Test		0.082	02:54
Air Blank		0.000	02:54
Subject Sample	# 1	0.143	02:55
Breath Volume		1.968	02:55
Air Blank		0.000	02:55
Air Blank		0.000	02:57
Subject Sample	# 2	0.136	02:58
Breath Volume		1.714	02:58
Air Blank		0.000	02:58
Control Test		0.081	02:59
Air Blank		0.000	02:59
Diagnostics Check		OK	02:59

Data Download Date/Time: 01/10/2020 12:00

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LESPERANCE,DANIEL,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CRUM,ZACHARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:34

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 12/19/2019 18:00:42

Diagnostics Check		OK	17:54
Air Blank		0.000	17:54
Control Test		0.081	17:54
Air Blank		0.000	17:55
Subject Sample	# 1	0.123	17:56
Breath Volume		1.417	17:56
Air Blank		0.000	17:56
Air Blank		0.000	17:58
Subject Sample	# 2	0.125	17:59
Breath Volume		1.386	17:59
Air Blank		0.000	17:59
Control Test		0.081	18:00
Air Blank		0.000	18:00
Diagnostics Check		OK	18:00

Data Download Date/Time: 12/23/2019 09:35

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,MATTHEW,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CARTER,MATTHEW
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 05:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: PARK

DATE/TIME OF TEST: 12/20/2019 05:26:12

Diagnostics Check		OK	05:19
Air Blank		0.000	05:20
Control Test		0.081	05:20
Air Blank		0.000	05:21
Subject Sample	# 1	0.155	05:21
Breath Volume		1.843	05:21
Air Blank		0.000	05:22
Air Blank		0.000	05:24
Subject Sample	# 2	0.151	05:24
Breath Volume		1.562	05:24
Air Blank		0.000	05:25
Control Test		0.081	05:25
Air Blank		0.000	05:26
Diagnostics Check		OK	05:26

Data Download Date/Time: 12/23/2019 09:35

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,MATTHEW,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CARTER,MATTHEW
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: DOWNEY

DATE/TIME OF TEST: 12/22/2019 04:19:31

Diagnostics Check		OK	04:12
Air Blank		0.000	04:12
Control Test		0.080	04:13
Air Blank		0.000	04:13
Subject Sample	# 1	0.191	04:14
Breath Volume		1.355	04:14
Air Blank		0.000	04:15
Air Blank		0.000	04:17
Subject Sample	# 2	0.188	04:18
Breath Volume		1.703	04:18
Air Blank		0.000	04:18
Control Test		0.080	04:18
Air Blank		0.000	04:19
Diagnostics Check		OK	04:19

Data Download Date/Time: 12/23/2019 09:35

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,MATTHEW,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ADAMS,DWAYNE
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: BARKER

DATE/TIME OF TEST: 12/22/2019 05:32:56

Diagnostics Check		OK	05:26
Air Blank		0.000	05:26
Control Test		0.081	05:26
Air Blank		0.000	05:27
Subject Sample	# 1	0.117	05:28
Breath Volume		1.750	05:28
Air Blank		0.000	05:28
Air Blank		0.000	05:30
Subject Sample	# 2	0.114	05:31
Breath Volume		1.191	05:31
Air Blank		0.000	05:32
Control Test		0.080	05:32
Air Blank		0.000	05:32
Diagnostics Check		OK	05:32

Data Download Date/Time: 12/23/2019 09:35

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOWRY,JAMIE,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SMITH,EMILY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:16

SUBJECT'S LAST NAME: RAY

DATE/TIME OF TEST: 12/22/2019 23:15:46

Diagnostics Check		OK	23:06	
Air Blank		0.000	23:06	
Control Test		0.081	23:07	
Air Blank		0.000	23:07	
Subject Sample	# 1	0.299	23:09	
Breath Volume		1.933	23:09	
Air Blank		0.000	23:10	
Air Blank		0.000	23:11	
Subject Sample	# 2		23:14	Subject Test Refused
Breath Volume		0.105	23:14	
Air Blank		0.000	23:14	
Control Test		0.078	23:15	
Air Blank		0.000	23:15	
Diagnostics Check		OK	23:15	

Data Download Date/Time: 12/23/2019 09:35

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLINE,NICOLE,S
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: COLLINS,RONALD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: BURGOS

DATE/TIME OF TEST: 12/27/2019 03:42:33

Diagnostics Check		OK	03:35
Air Blank		0.000	03:36
Control Test		0.080	03:36
Air Blank		0.000	03:37
Subject Sample	# 1	0.124	03:37
Breath Volume		1.441	03:37
Air Blank		0.000	03:38
Air Blank		0.000	03:40
Subject Sample	# 2	0.128	03:41
Breath Volume		2.050	03:41
Air Blank		0.000	03:41
Control Test		0.081	03:41
Air Blank		0.000	03:42
Diagnostics Check		OK	03:42

Data Download Date/Time: 01/22/2020 11:00

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAMOUR,ZACHARY,T
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CRUM,ZACKARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: RIGAUD

DATE/TIME OF TEST: 12/28/2019 03:01:28

Diagnostics Check		OK	02:58	
Air Blank		0.000	02:58	
Control Test		0.081	02:58	
Air Blank		0.000	02:59	
Subject Sample	# 1		03:00	Subject Test Refused
Breath Volume		0.000	03:00	
Air Blank		0.000	03:00	
Control Test		0.081	03:00	
Air Blank		0.000	03:01	
Diagnostics Check		OK	03:01	

Data Download Date/Time: 01/22/2020 11:00

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLINE,NICOLE,S
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ADAMS,DWAYNE
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 05:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:13

SUBJECT'S LAST NAME: GARDENHOUSE

DATE/TIME OF TEST: 12/28/2019 07:03:43

Diagnostics Check		OK	07:00	
Air Blank		0.000	07:01	
Control Test		0.081	07:01	
Air Blank		0.000	07:02	
Subject Sample	# 1		07:02	Subject Test Refused
Breath Volume		0.000	07:02	
Air Blank		0.000	07:02	
Control Test		0.081	07:03	
Air Blank		0.000	07:03	
Diagnostics Check		OK	07:03	

Data Download Date/Time: 01/22/2020 11:01

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 12/08/2019 03:36:18

Diagnostics Check		OK	03:29
Air Blank		0.000	03:29
Control Test		0.082	03:30
Air Blank		0.000	03:30
Subject Sample	# 1	0.128	03:31
Breath Volume		1.226	03:31
Air Blank		0.000	03:32
Air Blank		0.000	03:34
Subject Sample	# 2	0.139	03:34
Breath Volume		1.859	03:34
Air Blank		0.000	03:35
Control Test		0.080	03:35
Air Blank		0.000	03:36
Diagnostics Check		OK	03:36

Data Download Date/Time: 12/30/2019 12:57

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 19:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:56

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 12/12/2019 20:24:12

Diagnostics Check		OK	20:16
Air Blank		0.000	20:17
Control Test		0.081	20:17
Air Blank		0.000	20:18
Subject Sample	# 1	0.185	20:18
Breath Volume		1.656	20:18
Air Blank		0.000	20:19
Air Blank		0.000	20:21
Subject Sample	# 2	0.178	20:22
Breath Volume		1.273	20:22
Air Blank		0.000	20:23
Control Test		0.080	20:23
Air Blank		0.000	20:24
Diagnostics Check		OK	20:24

Data Download Date/Time: 12/30/2019 12:57

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,TIMOTHY,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: WHITE,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 20:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: BRAZIL

DATE/TIME OF TEST: 12/13/2019 21:25:30

Diagnostics Check		OK	21:13	
Air Blank		0.000	21:14	
Control Test		0.081	21:14	
Air Blank		0.000	21:14	
Subject Sample	# 1	0.233	21:18	Volume Not Met
Breath Volume		0.042	21:18	
Air Blank		0.000	21:18	
Air Blank		0.000	21:20	
Subject Sample	# 2	0.244	21:23	Volume Not Met
Breath Volume		0.441	21:23	
Air Blank		0.000	21:24	
Control Test		0.079	21:24	
Air Blank		0.000	21:25	
Diagnostics Check		OK	21:25	

Data Download Date/Time: 12/30/2019 12:57

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,TIMOTHY,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: WHITE,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 20:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: BRIZIL

DATE/TIME OF TEST: 12/13/2019 21:46:50

Diagnostics Check		OK	21:31	
Air Blank		0.000	21:31	
Control Test		0.079	21:31	
Air Blank		0.000	21:32	
Subject Sample	# 1	0.244	21:33	
Breath Volume		1.539	21:33	
Air Blank		0.000	21:34	
Air Blank		0.000	21:36	
Subject Sample	# 2	0.219	21:39	Volume Not Met
Breath Volume		0.257	21:39	
Air Blank		0.000	21:40	
Air Blank		0.000	21:42	
Subject Sample	# 3	0.219	21:45	Volume Not Met
Breath Volume		0.269	21:45	
Air Blank		0.000	21:45	
Control Test		0.076	21:46	
Air Blank		0.000	21:46	
Diagnostics Check		OK	21:46	

Data Download Date/Time: 12/30/2019 12:57

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENNESSEY,BRANDON,L
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: APONTE,ALEXANDRIA
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: WALTERS

DATE/TIME OF TEST: 12/14/2019 01:54:03

Diagnostics Check		OK	01:47
Air Blank		0.000	01:48
Control Test		0.080	01:48
Air Blank		0.000	01:48
Subject Sample	# 1	0.155	01:49
Breath Volume		1.851	01:49
Air Blank		0.000	01:50
Air Blank		0.000	01:51
Subject Sample	# 2	0.155	01:52
Breath Volume		1.410	01:52
Air Blank		0.000	01:53
Control Test		0.081	01:53
Air Blank		0.000	01:53
Diagnostics Check		OK	01:54

Data Download Date/Time: 12/30/2019 12:57

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: COLBARY

DATE/TIME OF TEST: 12/15/2019 04:30:48

Diagnostics Check		OK	04:23
Air Blank		0.000	04:24
Control Test		0.082	04:24
Air Blank		0.000	04:24
Subject Sample	# 1	0.204	04:25
Breath Volume		1.285	04:25
Air Blank		0.000	04:26
Air Blank		0.000	04:28
Subject Sample	# 2	0.200	04:29
Breath Volume		1.390	04:29
Air Blank		0.000	04:29
Control Test		0.081	04:30
Air Blank		0.000	04:30
Diagnostics Check		OK	04:30

Data Download Date/Time: 12/30/2019 12:57

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUSCIO,JESSE,J
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: BRIGGS,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 16:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:44

SUBJECT'S LAST NAME: RIX

DATE/TIME OF TEST: 12/19/2019 18:10:26

Diagnostics Check		OK	18:04
Air Blank		0.000	18:04
Control Test		0.083	18:05
Air Blank		0.000	18:05
Subject Sample	# 1	0.000	18:06
Breath Volume		3.894	18:06
Air Blank		0.000	18:06
Air Blank		0.000	18:08
Subject Sample	# 2	0.000	18:09
Breath Volume		3.542	18:09
Air Blank		0.000	18:09
Control Test		0.081	18:09
Air Blank		0.000	18:10
Diagnostics Check		OK	18:10

Data Download Date/Time: 12/30/2019 12:57

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YARBOROUGH,ERIC,L
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: KOTSIS,JOSHUA
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: MCLAUGHLIN

DATE/TIME OF TEST: 12/20/2019 01:25:02

Diagnostics Check		OK	01:18
Air Blank		0.000	01:19
Control Test		0.081	01:19
Air Blank		0.000	01:20
Subject Sample	# 1	0.088	01:20
Breath Volume		1.906	01:20
Air Blank		0.000	01:21
Air Blank		0.000	01:23
Subject Sample	# 2	0.080	01:23
Breath Volume		1.250	01:23
Air Blank		0.000	01:24
Control Test		0.081	01:24
Air Blank		0.000	01:24
Diagnostics Check		OK	01:25

Data Download Date/Time: 12/30/2019 12:57

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIMA,MICHAEL,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: LIMA,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: COELHO

DATE/TIME OF TEST: 12/21/2019 01:08:21

Diagnostics Check		OK	01:00
Air Blank		0.000	01:00
Control Test		0.082	01:00
Air Blank		0.000	01:01
Subject Sample	# 1	0.202	01:03
Breath Volume		1.335	01:03
Air Blank		0.000	01:03
Air Blank		0.000	01:05
Subject Sample	# 2	0.204	01:06
Breath Volume		1.183	01:06
Air Blank		0.000	01:07
Control Test		0.081	01:07
Air Blank		0.000	01:08
Diagnostics Check		OK	01:08

Data Download Date/Time: 12/30/2019 12:57

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 04:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: OBRIEN

DATE/TIME OF TEST: 12/22/2019 05:11:47

Diagnostics Check		OK	05:05
Air Blank		0.000	05:06
Control Test		0.081	05:06
Air Blank		0.000	05:06
Subject Sample	# 1	0.073	05:07
Breath Volume		2.359	05:07
Air Blank		0.000	05:08
Air Blank		0.000	05:09
Subject Sample	# 2	0.068	05:10
Breath Volume		2.179	05:10
Air Blank		0.000	05:10
Control Test		0.081	05:11
Air Blank		0.000	05:11
Diagnostics Check		OK	05:11

Data Download Date/Time: 12/30/2019 12:57

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: SPAHIU

DATE/TIME OF TEST: 12/23/2019 01:30:44

Diagnostics Check		OK	01:27	
Air Blank		0.000	01:28	
Control Test		0.081	01:28	
Air Blank		0.000	01:29	
Subject Sample	# 1		01:29	Subject Test Refused
Breath Volume		0.000	01:29	
Air Blank		0.000	01:29	
Control Test		0.082	01:30	
Air Blank		0.000	01:30	
Diagnostics Check		OK	01:30	

Data Download Date/Time: 12/30/2019 12:57

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YARBOROUGH,ERIC,L
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: ARGITIS,NICHOLAS
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 12/28/2019 03:22:35

Diagnostics Check		OK	03:16
Air Blank		0.000	03:16
Control Test		0.081	03:16
Air Blank		0.000	03:17
Subject Sample	# 1	0.138	03:18
Breath Volume		1.964	03:18
Air Blank		0.000	03:18
Air Blank		0.000	03:20
Subject Sample	# 2	0.138	03:21
Breath Volume		1.550	03:21
Air Blank		0.000	03:21
Control Test		0.080	03:22
Air Blank		0.000	03:22
Diagnostics Check		OK	03:22

Data Download Date/Time: 12/30/2019 12:57

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,MASON,E
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: THOMPSON,MASON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 05:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:25

SUBJECT'S LAST NAME: GERSDORF

DATE/TIME OF TEST: 12/28/2019 07:17:45

Diagnostics Check		OK	07:11
Air Blank		0.000	07:12
Control Test		0.081	07:12
Air Blank		0.000	07:13
Subject Sample	# 1	0.000	07:13
Breath Volume		1.562	07:13
Air Blank		0.000	07:14
Air Blank		0.000	07:16
Subject Sample	# 2	0.000	07:16
Breath Volume		1.531	07:16
Air Blank		0.000	07:16
Control Test		0.081	07:17
Air Blank		0.000	07:17
Diagnostics Check		OK	07:17

Data Download Date/Time: 12/30/2019 12:57

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORALUZZO,MICHAEL,G
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: CORALUZZO,MIKE
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: HUDER

DATE/TIME OF TEST: 12/29/2019 03:49:24

Diagnostics Check		OK	03:43	
Air Blank		0.000	03:44	
Control Test		0.079	03:44	
Air Blank		0.000	03:45	
Subject Sample	# 1	0.119	03:45	
Breath Volume		3.058	03:45	
Air Blank		0.000	03:46	
Air Blank		0.000	03:48	
Subject Sample	# 2	0.116	03:48	
Breath Volume		3.375	03:48	
Air Blank			03:48	RFI Detect
Air Blank		0.000	03:49	

Data Download Date/Time: 12/30/2019 12:57

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUSCIO,JESSE,J
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: BRIGGS,JASON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 15:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:04

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 12/29/2019 16:30:51

Diagnostics Check		OK	16:24
Air Blank		0.000	16:25
Control Test		0.081	16:25
Air Blank		0.000	16:25
Subject Sample	# 1	0.374	16:26
Breath Volume		2.746	16:26
Air Blank		0.000	16:27
Air Blank		0.000	16:28
Subject Sample	# 2	0.374	16:29
Breath Volume		2.015	16:29
Air Blank		0.000	16:29
Control Test		0.080	16:30
Air Blank		0.000	16:30
Diagnostics Check		OK	16:30

Data Download Date/Time: 12/30/2019 12:57

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYTHOLAR,RAYMOND,W
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: EVANS,IAN
ARRESTING OFFICER AGENCY: St. Augustine Beach
LAST AGENCY INSPECTION: 12/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: MYTHOLAR

DATE/TIME OF TEST: 12/30/2019 22:58:16

Diagnostics Check		OK	22:50
Air Blank		0.000	22:51
Control Test		0.081	22:51
Air Blank		0.000	22:51
Subject Sample	# 1	0.271	22:53
Breath Volume		2.355	22:53
Air Blank		0.000	22:54
Air Blank		0.000	22:56
Subject Sample	# 2	0.267	22:56
Breath Volume		2.289	22:56
Air Blank		0.000	22:57
Control Test		0.080	22:57
Air Blank		0.000	22:58
Diagnostics Check		OK	22:58

Data Download Date/Time: 01/29/2020 14:38

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001656

INSTRUMENT REGISTERED TO: NORTH MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PINEDA,MIGUEL,A
BREATH TEST OPERATOR AGENCY: NORTH MIAMI PD
ARRESTING OFFICER: PUPO,MICHAEL
ARRESTING OFFICER AGENCY: NORTH MIAMI PD
LAST AGENCY INSPECTION: 12/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34418080A3
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: ROSE

DATE/TIME OF TEST: 12/10/2019 00:27:42

Diagnostics Check		OK	00:24	
Air Blank		0.000	00:25	
Control Test		0.081	00:25	
Air Blank		0.000	00:25	
Subject Sample	# 1		00:26	Subject Test Refused
Breath Volume		0.000	00:26	
Air Blank		0.000	00:26	
Control Test		0.081	00:27	
Air Blank		0.000	00:27	
Diagnostics Check		OK	00:27	

Data Download Date/Time: 01/03/2020 07:51

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001656

INSTRUMENT REGISTERED TO: NORTH MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PINEDA,MIGUEL,A
BREATH TEST OPERATOR AGENCY: NORTH MIAMI PD
ARRESTING OFFICER: JONES,GARY
ARRESTING OFFICER AGENCY: NORTH MIAMI PD
LAST AGENCY INSPECTION: 12/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34418080A3
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: AGENOR

DATE/TIME OF TEST: 12/25/2019 05:22:26

Diagnostics Check		OK	05:15
Air Blank		0.000	05:15
Control Test		0.082	05:16
Air Blank		0.000	05:16
Subject Sample	# 1	0.184	05:17
Breath Volume		2.992	05:17
Air Blank		0.000	05:18
Air Blank		0.000	05:19
Subject Sample	# 2	0.180	05:20
Breath Volume		1.582	05:20
Air Blank		0.000	05:21
Control Test		0.081	05:21
Air Blank		0.000	05:22
Diagnostics Check		OK	05:22

Data Download Date/Time: 01/03/2020 07:51

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.