Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEADOWS,SHAWN,M
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: SMITH,SYDNEE
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 12:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:14

SUBJECT'S LAST NAME: RAMEY

DATE/TIME OF TEST: 12/10/2019 13:47:45

Diagnostics Check		OK	13:41
Air Blank		0.000	13:42
Control Test		0.079	13:42
Air Blank		0.000	13:42
Subject Sample	# 1	0.000	13:43
Breath Volume		3.363	13:43
Air Blank		0.000	13:43
Air Blank		0.000	13:45
Subject Sample	# 2	0.000	13:46
Breath Volume		3.488	13:46
Air Blank		0.000	13:46
Control Test		0.080	13:47
Air Blank		0.000	13:47
Diagnostics Check		OK	13:47

Data Download Date/Time: 12/30/2019 22:31

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIDDLE,JAMES,A
BREATH TEST OPERATOR AGENCY: Lake Mary PD
ARRESTING OFFICER: PIVOWAR,MICHAEL
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 13:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:20

SUBJECT'S LAST NAME: MCGILLIS

DATE/TIME OF TEST: 12/22/2019 14:59:20

Diagnostics Check		OK	14:44	
Air Blank		0.000	14:44	
Control Test		0.078	14:45	
Air Blank		0.000	14:45	
Subject Sample	# 1		14:48	No Sample Provided
Breath Volume		0.000	14:48	
Air Blank		0.000	14:49	
Air Blank		0.000	14:51	
Subject Sample	# 2	0.177	14:52	
Breath Volume		2.191	14:52	
Air Blank		0.000	14:52	
Air Blank		0.000	14:54	
Subject Sample	# 3		14:57	No Sample Provided
Breath Volume		0.000	14:57	
Air Blank		0.000	14:58	
Control Test		0.078	14:58	
Air Blank		0.000	14:59	
Diagnostics Check		OK	14:59	

Data Download Date/Time: 12/30/2019 22:31

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: HOLLAND,TROY
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 12/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 12/30/2019 04:12:03

Diagnostics Check		OK	04:05
Air Blank		0.000	04:06
Control Test		0.078	04:06
Air Blank		0.000	04:07
Subject Sample	# 1	0.119	04:07
Breath Volume		1.363	04:07
Air Blank		0.000	04:08
Air Blank		0.000	04:10
Subject Sample	# 2	0.119	04:10
Breath Volume		1.656	04:10
Air Blank		0.000	04:11
Control Test		0.079	04:11
Air Blank		0.000	04:11
Diagnostics Check		OK	04:12

Data Download Date/Time: 12/30/2019 22:31

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGANN,JAMES,D
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: MCGANN,JAMES
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 11/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: WAGNER

DATE/TIME OF TEST: 12/08/2019 02:11:01

Diagnostics Check		OK	02:03
Air Blank		0.000	02:04
Control Test		0.077	02:04
Air Blank		0.000	02:04
Subject Sample	# 1	0.199	02:05
Breath Volume		1.601	02:05
Air Blank		0.000	02:06
Air Blank		0.000	02:08
Subject Sample	# 2	0.200	02:09
Breath Volume		1.562	02:09
Air Blank		0.000	02:10
Control Test		0.076	02:10
Air Blank		0.000	02:10
Diagnostics Check		OK	02:11

Data Download Date/Time: 01/29/2020 09:37

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOSKINS,JAMIE,L
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: ALVAREZ,OCTAVIO
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 11/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 22:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: CHAVEZMEDINA

DATE/TIME OF TEST: 12/12/2019 23:01:12

Diagnostics Check		OK	22:54
Air Blank		0.000	22:55
Control Test		0.079	22:55
Air Blank		0.000	22:56
Subject Sample	# 1	0.130	22:56
Breath Volume		4.246	22:56
Air Blank		0.000	22:57
Air Blank		0.000	22:59
Subject Sample	# 2	0.134	22:59
Breath Volume		4.351	22:59
Air Blank		0.000	23:00
Control Test		0.079	23:00
Air Blank		0.000	23:01
Diagnostics Check		OK	23:01

Data Download Date/Time: 12/31/2019 17:45

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERGER,CHARLES,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BERGER,CHARLES
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: UZCATEGUI

DATE/TIME OF TEST: 12/01/2019 02:13:36

Diagnostics Check		OK	02:06
Air Blank		0.000	02:07
Control Test		0.082	02:07
Air Blank		0.000	02:07
Subject Sample	# 1	0.079	02:09
Breath Volume		1.890	02:09
Air Blank		0.000	02:09
Air Blank		0.000	02:11
Subject Sample	# 2	0.078	02:12
Breath Volume		1.898	02:12
Air Blank		0.000	02:12
Control Test		0.081	02:13
Air Blank		0.000	02:13
Diagnostics Check		OK	02:13

Data Download Date/Time: 12/21/2019 21:30

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: SCHMIDT

DATE/TIME OF TEST: 12/01/2019 02:37:32

Diagnostics Check		OK	02:34	
Air Blank		0.000	02:35	
Control Test		0.082	02:35	
Air Blank		0.000	02:35	
Subject Sample	# 1		02:36	Subject Test Refused
Breath Volume		0.000	02:36	
Air Blank		0.000	02:36	
Control Test		0.082	02:36	
Air Blank		0.000	02:37	
Diagnostics Check		OK	02:37	

Data Download Date/Time: 12/21/2019 21:30

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANWORP,JUDSON,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: VANWORP,JUDSON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: LOPEZ-EUCEDA

DATE/TIME OF TEST: 12/01/2019 04:39:02

Diagnostics Check		OK	04:31
Air Blank		0.000	04:32
Control Test		0.081	04:32
Air Blank		0.000	04:33
Subject Sample	# 1	0.186	04:34
Breath Volume		1.398	04:34
Air Blank		0.000	04:34
Air Blank		0.000	04:36
Subject Sample	# 2	0.187	04:37
Breath Volume		1.683	04:37
Air Blank		0.000	04:38
Control Test		0.081	04:38
Air Blank		0.000	04:38
Diagnostics Check		OK	04:39

Data Download Date/Time: 12/21/2019 21:30

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNAVELY,KENT,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HUDDLESTON,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 14:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:10

SUBJECT'S LAST NAME: MCGRAW

DATE/TIME OF TEST: 12/01/2019 15:37:16

Diagnostics Check		OK	15:30
Air Blank		0.000	15:30
Control Test		0.082	15:30
Air Blank		0.000	15:31
Subject Sample	# 1	0.181	15:32
Breath Volume		2.179	15:32
Air Blank		0.000	15:33
Air Blank		0.000	15:35
Subject Sample	# 2	0.180	15:35
Breath Volume		1.113	15:35
Air Blank		0.000	15:36
Control Test		0.081	15:36
Air Blank		0.000	15:37
Diagnostics Check		OK	15:37

Data Download Date/Time: 12/21/2019 21:30

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONGMIRE,BENJAMIN,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: LONGMIRE,BENJAMIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 18:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:48

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 12/01/2019 19:18:47

Diagnostics Check		OK	19:09
Air Blank		0.000	19:09
Control Test		0.081	19:10
Air Blank		0.000	19:10
Subject Sample	# 1	0.244	19:11
Breath Volume		1.636	19:11
Air Blank		0.000	19:12
Air Blank		0.000	19:14
Subject Sample	# 2	0.228	19:17
Breath Volume		1.710	19:17
Air Blank		0.000	19:17
Control Test		0.082	19:18
Air Blank		0.000	19:18
Diagnostics Check		OK	19:18

Data Download Date/Time: 12/21/2019 21:31

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERGER,CHARLES,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BERGER,CHARLES
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: BARE

DATE/TIME OF TEST: 12/05/2019 00:09:53

Diagnostics Check		OK	00:02
Air Blank		0.000	00:03
Control Test		0.081	00:03
Air Blank		0.000	00:03
Subject Sample	# 1	0.318	00:05
Breath Volume		1.988	00:05
Air Blank		0.000	00:05
Air Blank		0.000	00:07
Subject Sample	# 2	0.309	00:08
Breath Volume		2.828	00:08
Air Blank		0.000	00:08
Control Test		0.080	00:09
Air Blank		0.000	00:09
Diagnostics Check		OK	00:09

Data Download Date/Time: 12/21/2019 21:31

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,WILLIAM,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: SHAFFER

DATE/TIME OF TEST: 12/05/2019 00:40:05

Diagnostics Check		OK	00:34
Air Blank		0.000	00:34
Control Test		0.081	00:34
Air Blank		0.000	00:35
Subject Sample	# 1	0.000	00:35
Breath Volume		2.257	00:35
Air Blank		0.000	00:36
Air Blank		0.000	00:38
Subject Sample	# 2	0.000	00:38
Breath Volume		1.402	00:38
Air Blank		0.000	00:39
Control Test		0.081	00:39
Air Blank		0.000	00:39
Diagnostics Check		OK	00:40

Data Download Date/Time: 12/21/2019 21:31

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANWORP,JUDSON,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: VANWORP,JUDSON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: STEPONGZI

DATE/TIME OF TEST: 12/05/2019 04:40:26

Diagnostics Check		OK	04:37	
Air Blank		0.000	04:38	
Control Test		0.081	04:38	
Air Blank		0.000	04:38	
Subject Sample	# 1		04:39	Subject Test Refused
Breath Volume		0.000	04:39	
Air Blank		0.000	04:39	
Control Test		0.081	04:39	
Air Blank		0.000	04:40	
Diagnostics Check		OK	04:40	

Data Download Date/Time: 12/21/2019 21:31

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICOTTILLI,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI,JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: LYCAN

DATE/TIME OF TEST: 12/06/2019 00:11:46

Diagnostics Check		OK	00:03
Air Blank		0.000	00:03
Control Test		0.082	00:04
Air Blank		0.000	00:04
Subject Sample	# 1	0.186	00:06
Breath Volume		1.359	00:06
Air Blank		0.000	00:07
Air Blank		0.000	00:09
Subject Sample	# 2	0.188	00:10
Breath Volume		1.398	00:10
Air Blank		0.000	00:10
Control Test		0.082	00:11
Air Blank		0.000	00:11
Diagnostics Check		OK	00:11

Data Download Date/Time: 12/21/2019 21:31

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERGER,CHARLES,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BERGER,CHARLES
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 18:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: BOGGS

DATE/TIME OF TEST: 12/06/2019 19:18:26

Diagnostics Check		OK	19:15	
Air Blank		0.000	19:15	
Control Test		0.082	19:16	
Air Blank		0.000	19:16	
Subject Sample	# 1		19:17	Subject Test Refused
Breath Volume		0.000	19:17	
Air Blank		0.000	19:17	
Control Test		0.082	19:17	
Air Blank		0.000	19:18	
Diagnostics Check		OK	19:18	

Data Download Date/Time: 12/21/2019 21:31

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: BRINA

DATE/TIME OF TEST: 12/06/2019 22:06:49

Diagnostics Check		OK	22:00
Air Blank		0.000	22:01
Control Test		0.081	22:01
Air Blank		0.000	22:01
Subject Sample	# 1	0.101	22:02
Breath Volume		1.988	22:02
Air Blank		0.000	22:03
Air Blank		0.000	22:04
Subject Sample	# 2	0.088	22:05
Breath Volume		1.156	22:05
Air Blank		0.000	22:05
Control Test		0.081	22:06
Air Blank		0.000	22:06
Diagnostics Check		OK	22:06

Data Download Date/Time: 12/21/2019 21:31

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANWORP,JUDSON,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: VANWORP,JUDSON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:43

SUBJECT'S LAST NAME: SYLVESTRE

DATE/TIME OF TEST: 12/06/2019 23:10:16

Diagnostics Check		OK	23:03
Air Blank		0.000	23:03
Control Test		0.082	23:04
Air Blank		0.000	23:04
Subject Sample	# 1	0.137	23:05
Breath Volume		2.589	23:05
Air Blank		0.000	23:06
Air Blank		0.000	23:08
Subject Sample	# 2	0.125	23:08
Breath Volume		1.257	23:08
Air Blank		0.000	23:09
Control Test		0.082	23:09
Air Blank		0.000	23:10
Diagnostics Check		OK	23:10

Data Download Date/Time: 12/21/2019 21:31

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: HARDWICK

DATE/TIME OF TEST: 12/07/2019 04:33:48

Diagnostics Check		OK	04:30	
Air Blank		0.000	04:31	
Control Test		0.082	04:31	
Air Blank		0.000	04:31	
Subject Sample	# 1		04:32	Subject Test Refused
Breath Volume		0.000	04:32	
Air Blank		0.000	04:32	
Control Test		0.081	04:33	
Air Blank		0.000	04:33	
Diagnostics Check		OK	04:33	

Data Download Date/Time: 12/21/2019 21:31

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,TABARIE,G
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SULLIVAN,TABARIE
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 06:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:38

SUBJECT'S LAST NAME: GLISINSKI

DATE/TIME OF TEST: 12/07/2019 08:05:15

Diagnostics Check		OK	07:58
Air Blank		0.000	07:59
Control Test		0.082	07:59
Air Blank		0.000	08:00
Subject Sample	# 1	0.155	08:00
Breath Volume		2.175	08:00
Air Blank		0.000	08:01
Air Blank		0.000	08:03
Subject Sample	# 2	0.148	08:03
Breath Volume		1.507	08:03
Air Blank		0.000	08:04
Control Test		0.081	08:04
Air Blank		0.000	08:05
Diagnostics Check		OK	08:05

Data Download Date/Time: 12/21/2019 21:31

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: DIVITO

DATE/TIME OF TEST: 12/08/2019 04:23:13

Diagnostics Check		OK	04:16
Air Blank		0.000	04:17
Control Test		0.081	04:17
Air Blank		0.000	04:18
Subject Sample	# 1	0.094	04:18
Breath Volume		1.402	04:18
Air Blank		0.000	04:19
Air Blank		0.000	04:21
Subject Sample	# 2	0.087	04:21
Breath Volume		1.312	04:21
Air Blank		0.000	04:22
Control Test		0.082	04:22
Air Blank		0.000	04:23
Diagnostics Check		OK	04:23

Data Download Date/Time: 12/21/2019 21:31

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,WILLIAM,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:17

SUBJECT'S LAST NAME: JUNOT

DATE/TIME OF TEST: 12/08/2019 05:45:00

Diagnostics Check		OK	05:38
Air Blank		0.000	05:38
Control Test		0.081	05:39
Air Blank		0.000	05:39
Subject Sample	# 1	0.126	05:40
Breath Volume		2.460	05:40
Air Blank		0.000	05:40
Air Blank		0.000	05:42
Subject Sample	# 2	0.123	05:43
Breath Volume		1.859	05:43
Air Blank		0.000	05:44
Control Test		0.081	05:44
Air Blank		0.000	05:44
Diagnostics Check		OK	05:45

Data Download Date/Time: 12/21/2019 21:31

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ARBANAS,TROOPER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 15:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:30

SUBJECT'S LAST NAME: FORD

DATE/TIME OF TEST: 12/08/2019 16:41:10

Diagnostics Check		OK	16:34
Air Blank		0.000	16:35
Control Test		0.081	16:35
Air Blank		0.000	16:36
Subject Sample	# 1	0.172	16:36
Breath Volume		2.046	16:36
Air Blank		0.000	16:37
Air Blank		0.000	16:39
Subject Sample	# 2	0.182	16:39
Breath Volume		1.906	16:39
Air Blank		0.000	16:40
Control Test		0.081	16:40
Air Blank		0.000	16:41
Diagnostics Check		OK	16:41

Data Download Date/Time: 12/21/2019 21:31

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUGO-FELIX,DEPUTY,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PEREZ,DEPUTY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 16:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:02

SUBJECT'S LAST NAME: HAFKE

DATE/TIME OF TEST: 12/09/2019 17:37:54

Diagnostics Check		OK	17:31
Air Blank		0.000	17:31
Control Test		0.081	17:32
Air Blank		0.000	17:32
Subject Sample	# 1	0.119	17:33
Breath Volume		1.652	17:33
Air Blank		0.000	17:33
Air Blank		0.000	17:35
Subject Sample	# 2	0.117	17:36
Breath Volume		1.945	17:36
Air Blank		0.000	17:37
Control Test		0.081	17:37
Air Blank		0.000	17:37
Diagnostics Check		OK	17:37

Data Download Date/Time: 12/21/2019 21:31

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE,DEPUTY WILTSE,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WILTSE,DEPUTY WILLIAM
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: SOUTER

DATE/TIME OF TEST: 12/11/2019 02:53:12

Diagnostics Check		OK	02:50	
Air Blank		0.000	02:50	
Control Test		0.081	02:50	
Air Blank		0.000	02:51	
Subject Sample	# 1		02:51	Subject Test Refused
Breath Volume		0.000	02:51	
Air Blank		0.000	02:52	
Control Test		0.081	02:52	
Air Blank		0.000	02:53	
Diagnostics Check		OK	02:53	

Data Download Date/Time: 12/21/2019 21:31

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICOTTILLI,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI,JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: CALLEROS

DATE/TIME OF TEST: 12/12/2019 03:15:57

Diagnostics Check		OK	03:10	
Air Blank		0.000	03:10	
Control Test		0.082	03:11	
Air Blank		0.000	03:11	
Subject Sample	# 1		03:14	Subject Test Refused
Breath Volume		0.000	03:14	
Air Blank		0.000	03:15	
Control Test		0.082	03:15	
Air Blank		0.000	03:15	
Diagnostics Check		OK	03:15	

Data Download Date/Time: 12/21/2019 21:31

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: GONZALEZ SANTIAGO

DATE/TIME OF TEST: 12/12/2019 04:26:14

Diagnostics Check		OK	04:18
Air Blank		0.000	04:18
Control Test		0.082	04:19
Air Blank		0.000	04:19
Subject Sample	# 1	0.187	04:21
Breath Volume		1.367	04:21
Air Blank		0.000	04:22
Air Blank		0.000	04:24
Subject Sample	# 2	0.194	04:24
Breath Volume		1.628	04:24
Air Blank		0.000	04:25
Control Test		0.080	04:25
Air Blank		0.000	04:26
Diagnostics Check		OK	04:26

Data Download Date/Time: 12/21/2019 21:31

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERGER,CHARLES,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BERGER,CHARLES
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: SIEBERT

DATE/TIME OF TEST: 12/12/2019 21:46:46

Diagnostics Check		OK	21:40
Air Blank		0.000	21:41
Control Test		0.081	21:41
Air Blank		0.000	21:41
Subject Sample	# 1	0.000	21:42
Breath Volume		3.531	21:42
Air Blank		0.000	21:42
Air Blank		0.000	21:44
Subject Sample	# 2	0.000	21:45
Breath Volume		2.023	21:45
Air Blank		0.000	21:45
Control Test		0.081	21:46
Air Blank		0.000	21:46
Diagnostics Check		OK	21:46

Data Download Date/Time: 12/21/2019 21:31

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 04:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: ORTEGA JAIME

DATE/TIME OF TEST: 12/13/2019 04:47:41

Diagnostics Check		OK	04:44	
Air Blank		0.000	04:45	
Control Test		0.081	04:45	
Air Blank		0.000	04:45	
Subject Sample	# 1		04:46	Subject Test Refused
Breath Volume		0.000	04:46	
Air Blank		0.000	04:46	
Control Test		0.081	04:47	
Air Blank		0.000	04:47	
Diagnostics Check		OK	04:47	

Data Download Date/Time: 12/21/2019 21:31

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERGER,CHARLES,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BERGER,CHARLES
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 18:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: JIMENEZ AGUILAR

DATE/TIME OF TEST: 12/13/2019 19:24:51

Diagnostics Check		OK	19:21	
Air Blank		0.000	19:21	
Control Test		0.081	19:22	
Air Blank		0.000	19:22	
Subject Sample	# 1		19:23	Subject Test Refused
Breath Volume		0.000	19:23	
Air Blank		0.000	19:23	
Control Test		0.082	19:24	
Air Blank		0.000	19:24	
Diagnostics Check		OK	19:24	

Data Download Date/Time: 12/21/2019 21:31

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICOTTILLI,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI,JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:06

SUBJECT'S LAST NAME: DRAUGHN

DATE/TIME OF TEST: 12/13/2019 22:35:16

Diagnostics Check		OK	22:27
Air Blank		0.000	22:28
Control Test		0.081	22:28
Air Blank		0.000	22:28
Subject Sample	# 1	0.043	22:30
Breath Volume		1.449	22:30
Air Blank		0.000	22:30
Air Blank		0.000	22:32
Subject Sample	# 2	0.040	22:33
Breath Volume		1.859	22:33
Air Blank		0.000	22:34
Control Test		0.082	22:34
Air Blank		0.000	22:35
Diagnostics Check		OK	22:35

Data Download Date/Time: 12/21/2019 21:31

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERGER,CHARLES,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BERGER,CHARLES
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: COLE

DATE/TIME OF TEST: 12/13/2019 23:47:10

Diagnostics Check		OK	23:40
Air Blank		0.000	23:40
Control Test		0.081	23:41
Air Blank		0.000	23:41
Subject Sample	# 1	0.228	23:42
Breath Volume		3.144	23:42
Air Blank		0.000	23:43
Air Blank		0.000	23:45
Subject Sample	# 2	0.233	23:45
Breath Volume		1.812	23:45
Air Blank		0.000	23:46
Control Test		0.081	23:46
Air Blank		0.000	23:47
Diagnostics Check		OK	23:47

Data Download Date/Time: 12/21/2019 21:31

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANWORP,JUDSON,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: VANWORP,JUDSON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: DUFFY

DATE/TIME OF TEST: 12/14/2019 01:17:54

Diagnostics Check		OK	01:12
Air Blank		0.000	01:12
Control Test		0.082	01:12
Air Blank		0.000	01:13
Subject Sample	# 1	0.000	01:13
Breath Volume		2.281	01:13
Air Blank		0.000	01:14
Air Blank		0.000	01:16
Subject Sample	# 2	0.000	01:16
Breath Volume		1.976	01:16
Air Blank		0.000	01:17
Control Test		0.081	01:17
Air Blank		0.000	01:17
Diagnostics Check		OK	01:17

Data Download Date/Time: 12/21/2019 21:31

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICOTTILLI,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI,JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: OLMEDA

DATE/TIME OF TEST: 12/14/2019 02:09:51

Diagnostics Check		OK	02:02
Air Blank		0.000	02:03
Control Test		0.081	02:03
Air Blank		0.000	02:04
Subject Sample	# 1	0.147	02:05
Breath Volume		1.304	02:05
Air Blank		0.000	02:05
Air Blank		0.000	02:07
Subject Sample	# 2	0.155	02:08
Breath Volume		1.453	02:08
Air Blank		0.000	02:08
Control Test		0.081	02:09
Air Blank		0.000	02:09
Diagnostics Check		OK	02:09

Data Download Date/Time: 12/21/2019 21:31

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANWORP,JUDSON,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: VANWORP,JUDSON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: KONTOS

DATE/TIME OF TEST: 12/14/2019 04:52:38

Diagnostics Check		OK	04:46
Air Blank		0.000	04:46
Control Test		0.081	04:46
Air Blank		0.000	04:47
Subject Sample	# 1	0.112	04:47
Breath Volume		2.011	04:47
Air Blank		0.000	04:48
Air Blank		0.000	04:50
Subject Sample	# 2	0.110	04:51
Breath Volume		1.507	04:51
Air Blank		0.000	04:51
Control Test		0.080	04:52
Air Blank		0.000	04:52
Diagnostics Check		OK	04:52

Data Download Date/Time: 12/21/2019 21:31

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TSANAKALIOTIS,REGINA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: CRANDALL

DATE/TIME OF TEST: 12/15/2019 23:47:20

Diagnostics Check		OK	23:41
Air Blank		0.000	23:41
Control Test		0.081	23:41
Air Blank		0.000	23:42
Subject Sample	# 1	0.282	23:42
Breath Volume		2.996	23:42
Air Blank		0.000	23:43
Air Blank		0.000	23:45
Subject Sample	# 2	0.278	23:45
Breath Volume		2.429	23:45
Air Blank		0.000	23:46
Control Test		0.081	23:46
Air Blank		0.000	23:47
Diagnostics Check		OK	23:47

Data Download Date/Time: 12/21/2019 21:31

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FITZGERALD,MARQUIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 12/16/2019 01:10:11

Diagnostics Check		OK	01:07	
Air Blank		0.000	01:07	
Control Test		0.080	01:08	
Air Blank		0.000	01:08	
Subject Sample	# 1		01:08	Subject Test Refused
Breath Volume		0.000	01:08	
Air Blank		0.000	01:09	
Control Test		0.081	01:09	
Air Blank		0.000	01:10	
Diagnostics Check		OK	01:10	

Data Download Date/Time: 12/21/2019 21:31

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: JOHNSON,MEGAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: GILMAN

DATE/TIME OF TEST: 12/18/2019 01:17:16

Diagnostics Check		OK	01:14	
Air Blank		0.000	01:14	
Control Test		0.082	01:15	
Air Blank		0.000	01:15	
Subject Sample	# 1		01:15	Subject Test Refused
Breath Volume		0.000	01:15	
Air Blank		0.000	01:16	
Control Test		0.081	01:16	
Air Blank		0.000	01:17	
Diagnostics Check		OK	01:17	

Data Download Date/Time: 12/21/2019 21:31

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNAVELY,KENT,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SNAVELY,KENT
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 10:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:00

SUBJECT'S LAST NAME: ARSENAULT

DATE/TIME OF TEST: 12/18/2019 11:23:18

Diagnostics Check		OK	11:20	
Air Blank		0.000	11:20	
Control Test		0.081	11:21	
Air Blank		0.000	11:21	
Subject Sample	# 1		11:21	Subject Test Refused
Breath Volume		0.000	11:21	
Air Blank		0.000	11:22	
Control Test		0.081	11:22	
Air Blank		0.000	11:23	
Diagnostics Check		OK	11:23	

Data Download Date/Time: 12/21/2019 21:31

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE,DEPUTY WILLIAM,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WILTSE,DEPUTY WILLIAM
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: KLEFEKER-THOMPSON

DATE/TIME OF TEST: 12/19/2019 04:20:33

Diagnostics Check		OK	04:17	
Air Blank		0.000	04:18	
Control Test		0.082	04:18	
Air Blank		0.000	04:18	
Subject Sample	# 1		04:19	Subject Test Refused
Breath Volume		0.000	04:19	
Air Blank		0.000	04:19	
Control Test		0.082	04:19	
Air Blank		0.000	04:20	
Diagnostics Check		OK	04:20	

Data Download Date/Time: 12/21/2019 21:31

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EASTTY,MARK,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: DAVIDSON

DATE/TIME OF TEST: 12/20/2019 21:22:07

Diagnostics Check		OK	21:16
Air Blank		0.000	21:16
Control Test		0.082	21:17
Air Blank		0.000	21:17
Subject Sample	# 1	0.000	21:17
Breath Volume		1.941	21:17
Air Blank		0.000	21:18
Air Blank		0.000	21:20
Subject Sample	# 2	0.000	21:20
Breath Volume		2.300	21:20
Air Blank		0.000	21:21
Control Test		0.082	21:21
Air Blank		0.000	21:22
Diagnostics Check		OK	21:22

Data Download Date/Time: 12/21/2019 21:31

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EASTTY,MARK,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TORO,JARED
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 12/20/2019 22:49:02

Diagnostics Check		OK	22:43
Air Blank		0.000	22:43
Control Test		0.082	22:43
Air Blank		0.000	22:44
Subject Sample	# 1	0.000	22:44
Breath Volume		2.203	22:44
Air Blank		0.000	22:45
Air Blank		0.000	22:47
Subject Sample	# 2	0.000	22:47
Breath Volume		1.953	22:47
Air Blank		0.000	22:48
Control Test		0.081	22:48
Air Blank		0.000	22:48
Diagnostics Check		OK	22:49

Data Download Date/Time: 12/21/2019 21:31

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EASTTY,MARK,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: VANWORP,JUDSON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 12/21/2019 00:09:14

Diagnostics Check		OK	00:03
Air Blank		0.000	00:03
Control Test		0.082	00:04
Air Blank		0.000	00:04
Subject Sample	# 1	0.000	00:05
Breath Volume		2.894	00:05
Air Blank		0.000	00:05
Air Blank		0.000	00:07
Subject Sample	# 2	0.000	00:07
Breath Volume		2.820	00:07
Air Blank		0.000	00:08
Control Test		0.081	00:08
Air Blank		0.000	00:09
Diagnostics Check		OK	00:09

Data Download Date/Time: 12/21/2019 21:31

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EASTTY,MARK,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: JOHNSON,MEGAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: GRIFFITH

DATE/TIME OF TEST: 12/21/2019 00:55:16

Diagnostics Check		OK	00:46	
Air Blank		0.000	00:47	
Control Test		0.082	00:47	
Air Blank		0.000	00:48	
Subject Sample	# 1		00:51	No Sample Provided
Breath Volume		0.000	00:51	
Air Blank		0.000	00:51	
Air Blank		0.000	00:53	
Subject Sample	# 2		00:53	Subject Test Refused
Breath Volume		0.000	00:53	
Air Blank		0.000	00:54	
Control Test		0.082	00:54	
Air Blank		0.000	00:55	
Diagnostics Check		OK	00:55	

Data Download Date/Time: 12/21/2019 21:31

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EASTTY,MARK,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: BENNINGFIELD

DATE/TIME OF TEST: 12/21/2019 01:31:29

Diagnostics Check		OK	01:25
Air Blank		0.000	01:25
Control Test		0.082	01:26
Air Blank		0.000	01:26
Subject Sample	# 1	0.070	01:27
Breath Volume		2.546	01:27
Air Blank		0.000	01:27
Air Blank		0.000	01:29
Subject Sample	# 2	0.069	01:29
Breath Volume		2.312	01:29
Air Blank		0.000	01:30
Control Test		0.081	01:30
Air Blank		0.000	01:31
Diagnostics Check		OK	01:31

Data Download Date/Time: 12/21/2019 21:31

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EASTTY,MARK,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI,JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: FUSSELL

DATE/TIME OF TEST: 12/21/2019 02:15:00

Diagnostics Check		OK	02:08
Air Blank		0.000	02:08
Control Test		0.082	02:08
Air Blank		0.000	02:09
Subject Sample	# 1	0.226	02:09
Breath Volume		1.683	02:09
Air Blank		0.000	02:10
Air Blank		0.000	02:12
Subject Sample	# 2	0.226	02:13
Breath Volume		2.148	02:13
Air Blank		0.000	02:14
Control Test		0.081	02:14
Air Blank		0.000	02:14
Diagnostics Check		OK	02:15

Data Download Date/Time: 12/21/2019 21:31

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EASTTY,MARK,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BOYLE,THOMAS
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: POWELL

DATE/TIME OF TEST: 12/21/2019 03:02:30

Diagnostics Check		OK	02:59	
Air Blank		0.000	03:00	
Control Test		0.081	03:00	
Air Blank		0.000	03:00	
Subject Sample	# 1		03:01	Subject Test Refused
Breath Volume		0.000	03:01	
Air Blank		0.000	03:01	
Control Test		0.082	03:01	
Air Blank		0.000	03:02	
Diagnostics Check		OK	03:02	

Data Download Date/Time: 12/21/2019 21:31

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EASTTY,MARK,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: YEARGIN,GENESIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: MYFTARAJ

DATE/TIME OF TEST: 12/21/2019 03:31:34

Diagnostics Check		OK	03:26	
Air Blank		0.000	03:27	
Control Test		0.081	03:27	
Air Blank		0.000	03:27	
Subject Sample	# 1		03:30	Subject Test Refused
Breath Volume		0.242	03:30	
Air Blank		0.000	03:30	
Control Test		0.082	03:30	
Air Blank		0.000	03:31	
Diagnostics Check		OK	03:31	

Data Download Date/Time: 12/21/2019 21:31

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNAVELY,KENT,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SHORT,JACLYN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 14:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:18

SUBJECT'S LAST NAME: BONSAL

DATE/TIME OF TEST: 12/24/2019 15:44:43

Diagnostics Check		OK	15:38
Air Blank		0.000	15:38
Control Test		0.081	15:39
Air Blank		0.000	15:39
Subject Sample	# 1	0.094	15:40
Breath Volume		1.652	15:40
Air Blank		0.000	15:40
Air Blank		0.000	15:42
Subject Sample	# 2	0.086	15:43
Breath Volume		1.812	15:43
Air Blank		0.000	15:43
Control Test		0.082	15:44
Air Blank		0.000	15:44
Diagnostics Check		OK	15:44

Data Download Date/Time: 01/11/2020 21:34

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICOTTILLI,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI,JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 17:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: PINE

DATE/TIME OF TEST: 12/27/2019 18:34:57

Diagnostics Check		OK	18:28
Air Blank		0.000	18:28
Control Test		0.081	18:28
Air Blank		0.000	18:29
Subject Sample	# 1	0.000	18:30
Breath Volume		1.488	18:30
Air Blank		0.000	18:31
Air Blank		0.000	18:33
Subject Sample	# 2	0.000	18:33
Breath Volume		1.554	18:33
Air Blank		0.000	18:34
Control Test		0.081	18:34
Air Blank		0.000	18:34
Diagnostics Check		OK	18:34

Data Download Date/Time: 01/11/2020 21:34

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNAVELY,KENT,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SNAVELY,KENT
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 07:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:10

SUBJECT'S LAST NAME: COLEMAN

DATE/TIME OF TEST: 12/28/2019 07:41:22

Diagnostics Check		OK	07:38	
Air Blank		0.000	07:38	
Control Test		0.081	07:39	
Air Blank		0.000	07:39	
Subject Sample	# 1		07:40	Subject Test Refused
Breath Volume		0.000	07:40	
Air Blank		0.000	07:40	
Control Test		0.082	07:40	
Air Blank		0.000	07:41	
Diagnostics Check		OK	07:41	

Data Download Date/Time: 01/11/2020 21:34

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 20:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 12/29/2019 21:16:37

Diagnostics Check		OK	21:10
Air Blank		0.000	21:11
Control Test		0.082	21:11
Air Blank		0.000	21:12
Subject Sample	# 1	0.000	21:12
Breath Volume		2.734	21:12
Air Blank		0.000	21:12
Air Blank		0.000	21:14
Subject Sample	# 2	0.000	21:15
Breath Volume		2.382	21:15
Air Blank		0.000	21:15
Control Test		0.081	21:16
Air Blank		0.000	21:16
Diagnostics Check		OK	21:16

Data Download Date/Time: 01/11/2020 21:34

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE,DEPUTY WILLIAM,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WILTSE,DEPUTY WILLIAM
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: LAESSER

DATE/TIME OF TEST: 12/30/2019 04:14:58

Diagnostics Check		OK	04:11	
Air Blank		0.000	04:12	
Control Test		0.083	04:12	
Air Blank		0.000	04:13	
Subject Sample	# 1		04:13	Subject Test Refused
Breath Volume		0.000	04:13	
Air Blank		0.000	04:14	
Control Test		0.082	04:14	
Air Blank		0.000	04:14	
Diagnostics Check		OK	04:14	

Data Download Date/Time: 01/11/2020 21:34

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROSS,M,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SKALKO,R
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: KNAPIK

DATE/TIME OF TEST: 12/02/2019 00:06:49

Diagnostics Check		OK	00:00
Air Blank		0.000	00:01
Control Test		0.080	00:01
Air Blank		0.000	00:01
Subject Sample	# 1	0.213	00:02
Breath Volume		2.914	00:02
Air Blank		0.000	00:03
Air Blank		0.000	00:04
Subject Sample	# 2	0.199	00:05
Breath Volume		1.414	00:05
Air Blank		0.000	00:05
Control Test		0.079	00:06
Air Blank		0.000	00:06
Diagnostics Check		OK	00:06

Data Download Date/Time: 12/21/2019 23:36

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROSS,M,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMIN,N
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: LOPEZ-HERNANDEZ

DATE/TIME OF TEST: 12/02/2019 02:42:39

Diagnostics Check		OK	02:37	
Air Blank		0.000	02:37	
Control Test		0.080	02:38	
Air Blank		0.000	02:38	
Subject Sample	# 1		02:40	Subject Test Refused
Breath Volume		0.347	02:40	
Air Blank		0.000	02:41	
Control Test		0.081	02:42	
Air Blank		0.000	02:42	
Diagnostics Check		OK	02:42	

Data Download Date/Time: 12/21/2019 23:36

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROSS,M,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DONOHUE,S
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: WIESBROOK

DATE/TIME OF TEST: 12/02/2019 03:10:45

Diagnostics Check		OK	03:07	
Air Blank		0.000	03:08	
Control Test		0.079	03:08	
Air Blank		0.000	03:09	
Subject Sample	# 1		03:09	Subject Test Refused
Breath Volume		0.000	03:09	
Air Blank		0.000	03:09	
Control Test		0.079	03:10	
Air Blank		0.000	03:10	
Diagnostics Check		OK	03:10	

Data Download Date/Time: 12/21/2019 23:36

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORTON,GREG,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HULME,JAMES
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:03

SUBJECT'S LAST NAME: DUBOIS

DATE/TIME OF TEST: 12/03/2019 17:30:14

Diagnostics Check		OK	17:23
Air Blank		0.000	17:24
Control Test		0.080	17:24
Air Blank		0.000	17:25
Subject Sample	# 1	0.000	17:25
Breath Volume		1.910	17:25
Air Blank		0.000	17:26
Air Blank		0.000	17:28
Subject Sample	# 2	0.000	17:28
Breath Volume		1.976	17:28
Air Blank		0.000	17:29
Control Test		0.081	17:29
Air Blank		0.000	17:30
Diagnostics Check		OK	17:30

Data Download Date/Time: 12/21/2019 23:36

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SALZMAN,DEPUTY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 12:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:32

SUBJECT'S LAST NAME: FRAZEE

DATE/TIME OF TEST: 12/07/2019 14:55:31

Diagnostics Check		OK	14:52	
Air Blank		0.000	14:52	
Control Test		0.080	14:53	
Air Blank		0.000	14:53	
Subject Sample	# 1		14:54	Subject Test Refused
Breath Volume		0.000	14:54	
Air Blank		0.000	14:54	
Control Test		0.080	14:54	
Air Blank		0.000	14:55	
Diagnostics Check		OK	14:55	

Data Download Date/Time: 12/21/2019 23:36

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICOTTILLI,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI,JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 12/08/2019 03:04:51

Diagnostics Check		OK	03:01	
Air Blank		0.000	03:02	
Control Test		0.079	03:02	
Air Blank		0.000	03:03	
Subject Sample	# 1		03:03	Subject Test Refused
Breath Volume		0.000	03:03	
Air Blank		0.000	03:03	
Control Test		0.080	03:04	
Air Blank		0.000	03:04	
Diagnostics Check		OK	03:04	

Data Download Date/Time: 12/21/2019 23:36

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SMITH,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: BRUSSELBACK

DATE/TIME OF TEST: 12/20/2019 22:35:00

Diagnostics Check		OK	22:28
Air Blank		0.000	22:28
Control Test		0.080	22:29
Air Blank		0.000	22:29
Subject Sample	# 1	0.126	22:30
Breath Volume		1.121	22:30
Air Blank		0.000	22:30
Air Blank		0.000	22:32
Subject Sample	# 2	0.127	22:33
Breath Volume		2.515	22:33
Air Blank		0.000	22:34
Control Test		0.080	22:34
Air Blank		0.000	22:34
Diagnostics Check		OK	22:35

Data Download Date/Time: 12/21/2019 23:36

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: FLETCHER

DATE/TIME OF TEST: 12/21/2019 00:00:17

Diagnostics Check		OK	23:57	
Air Blank		0.000	23:57	
Control Test		0.080	23:58	
Air Blank		0.000	23:58	
Subject Sample	# 1		23:58	Subject Test Refused
Breath Volume		0.000	23:58	
Air Blank		0.000	23:59	
Control Test		0.080	23:59	
Air Blank		0.000	00:00	
Diagnostics Check		OK	00:00	

Data Download Date/Time: 12/21/2019 23:36

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: REDMAN

DATE/TIME OF TEST: 12/21/2019 00:45:40

Diagnostics Check		OK	00:39
Air Blank		0.000	00:39
Control Test		0.081	00:40
Air Blank		0.000	00:40
Subject Sample	# 1	0.042	00:41
Breath Volume		2.367	00:41
Air Blank		0.000	00:41
Air Blank		0.000	00:43
Subject Sample	# 2	0.040	00:44
Breath Volume		2.136	00:44
Air Blank		0.000	00:44
Control Test		0.079	00:45
Air Blank		0.000	00:45
Diagnostics Check		OK	00:45

Data Download Date/Time: 12/21/2019 23:36

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: HUGHES

DATE/TIME OF TEST: 12/21/2019 01:17:33

Diagnostics Check		OK	01:09
Air Blank		0.000	01:10
Control Test		0.080	01:10
Air Blank		0.000	01:11
Subject Sample	# 1	0.081	01:12
Breath Volume		1.289	01:12
Air Blank		0.000	01:13
Air Blank		0.000	01:14
Subject Sample	# 2	0.082	01:15
Breath Volume		1.535	01:15
Air Blank		0.000	01:16
Control Test		0.079	01:16
Air Blank		0.000	01:17
Diagnostics Check		OK	01:17

Data Download Date/Time: 12/21/2019 23:36

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BERGER,CHARLES
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: ARGENZIO

DATE/TIME OF TEST: 12/21/2019 02:12:15

Diagnostics Check		OK	02:05
Air Blank		0.000	02:05
Control Test		0.079	02:05
Air Blank		0.000	02:06
Subject Sample	# 1	0.241	02:06
Breath Volume		3.117	02:06
Air Blank		0.000	02:07
Air Blank		0.000	02:09
Subject Sample	# 2	0.241	02:10
Breath Volume		3.046	02:10
Air Blank		0.000	02:11
Control Test		0.079	02:11
Air Blank		0.000	02:12
Diagnostics Check		OK	02:12

Data Download Date/Time: 12/21/2019 23:36

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HANSEN,DUSTIN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: BAGLAS

DATE/TIME OF TEST: 12/21/2019 03:22:37

Diagnostics Check		OK	03:19	
Air Blank		0.000	03:20	
Control Test		0.080	03:20	
Air Blank		0.000	03:21	
Subject Sample	# 1		03:21	Subject Test Refused
Breath Volume		0.000	03:21	
Air Blank		0.000	03:21	
Control Test		0.080	03:22	
Air Blank		0.000	03:22	
Diagnostics Check		OK	03:22	

Data Download Date/Time: 12/21/2019 23:36

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: CASPER

DATE/TIME OF TEST: 12/21/2019 04:18:43

Diagnostics Check		OK	04:12
Air Blank		0.000	04:12
Control Test		0.080	04:13
Air Blank		0.000	04:13
Subject Sample	# 1	0.193	04:13
Breath Volume		2.769	04:13
Air Blank		0.000	04:14
Air Blank		0.000	04:16
Subject Sample	# 2	0.191	04:17
Breath Volume		2.890	04:17
Air Blank		0.000	04:17
Control Test		0.080	04:18
Air Blank		0.000	04:18
Diagnostics Check		OK	04:18

Data Download Date/Time: 12/21/2019 23:36

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MITCHELL,ALEXANDRA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: CAMENS

DATE/TIME OF TEST: 12/21/2019 04:39:41

Diagnostics Check		OK	04:32
Air Blank		0.000	04:33
Control Test		0.080	04:33
Air Blank		0.000	04:33
Subject Sample	# 1	0.187	04:34
Breath Volume		1.531	04:34
Air Blank		0.000	04:35
Air Blank		0.000	04:36
Subject Sample	# 2	0.190	04:37
Breath Volume		1.964	04:37
Air Blank		0.000	04:38
Control Test		0.080	04:39
Air Blank		0.000	04:39
Diagnostics Check		OK	04:39

Data Download Date/Time: 12/21/2019 23:36

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROSS,M,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: JOWELL,T
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: SASKO

DATE/TIME OF TEST: 12/24/2019 23:04:54

Diagnostics Check		OK	22:58
Air Blank		0.000	22:59
Control Test		0.080	22:59
Air Blank		0.000	22:59
Subject Sample	# 1	0.115	23:00
Breath Volume		2.089	23:00
Air Blank		0.000	23:01
Air Blank		0.000	23:02
Subject Sample	# 2	0.113	23:03
Breath Volume		2.386	23:03
Air Blank		0.000	23:03
Control Test		0.080	23:04
Air Blank		0.000	23:04
Diagnostics Check		OK	23:04

Data Download Date/Time: 01/11/2020 23:18

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROSS,M,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MULLINS,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: HARTLEY

DATE/TIME OF TEST: 12/25/2019 04:04:56

Diagnostics Check		OK	03:58
Air Blank		0.000	03:59
Control Test		0.080	03:59
Air Blank		0.000	03:59
Subject Sample	# 1	0.083	04:00
Breath Volume		2.195	04:00
Air Blank		0.000	04:00
Air Blank		0.000	04:02
Subject Sample	# 2	0.082	04:03
Breath Volume		2.550	04:03
Air Blank		0.000	04:04
Control Test		0.080	04:04
Air Blank		0.000	04:04
Diagnostics Check		OK	04:04

Data Download Date/Time: 01/11/2020 23:18

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

INSTRUMENT REGISTERED TO: INDIALANTIC P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIBSON,ZACHARY,A
BREATH TEST OPERATOR AGENCY: Indialantic PD
ARRESTING OFFICER: THATCHER,MATTHEW
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 11/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: ROSELIP

DATE/TIME OF TEST: 12/01/2019 03:42:17

Diagnostics Check		OK	03:36
Air Blank		0.000	03:36
Control Test		0.079	03:36
Air Blank		0.000	03:37
Subject Sample	# 1	0.073	03:37
Breath Volume		3.710	03:37
Air Blank		0.000	03:38
Air Blank		0.000	03:40
Subject Sample	# 2	0.074	03:40
Breath Volume		3.691	03:40
Air Blank		0.000	03:41
Control Test		0.078	03:41
Air Blank		0.000	03:42
Diagnostics Check		OK	03:42

Data Download Date/Time: 12/17/2019 01:15

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

INSTRUMENT REGISTERED TO: INDIALANTIC P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: HALL,NICHOLAS
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 11/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: FRY

DATE/TIME OF TEST: 12/05/2019 23:04:25

Diagnostics Check		OK	22:58
Air Blank		0.000	22:58
Control Test		0.080	22:59
Air Blank		0.000	22:59
Subject Sample	# 1	0.148	22:59
Breath Volume		1.265	22:59
Air Blank		0.000	23:00
Air Blank		0.000	23:02
Subject Sample	# 2	0.149	23:02
Breath Volume		2.046	23:02
Air Blank		0.000	23:03
Control Test		0.081	23:03
Air Blank		0.000	23:04
Diagnostics Check		OK	23:04

Data Download Date/Time: 12/17/2019 01:15

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

INSTRUMENT REGISTERED TO: INDIALANTIC P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,WILLIAM,
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: HOWARD,WILLIAM
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 12/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: MUNIZ

DATE/TIME OF TEST: 12/19/2019 23:51:32

Diagnostics Check		OK	23:45
Air Blank		0.000	23:45
Control Test		0.080	23:45
Air Blank		0.000	23:46
Subject Sample	# 1	0.143	23:47
Breath Volume		1.625	23:47
Air Blank		0.000	23:47
Air Blank		0.000	23:49
Subject Sample	# 2	0.139	23:50
Breath Volume		1.488	23:50
Air Blank		0.000	23:50
Control Test		0.080	23:50
Air Blank		0.000	23:51
Diagnostics Check		OK	23:51

Data Download Date/Time: 01/28/2020 13:21

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

INSTRUMENT REGISTERED TO: INDIALANTIC P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEAL,ADAM,M
BREATH TEST OPERATOR AGENCY: Indialantic PD
ARRESTING OFFICER: BEAL,ADAM
ARRESTING OFFICER AGENCY: Indialantic PD
LAST AGENCY INSPECTION: 12/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: BECKER

DATE/TIME OF TEST: 12/20/2019 01:45:13

Diagnostics Check		OK	01:39
Air Blank		0.000	01:39
Control Test		0.080	01:39
Air Blank		0.000	01:40
Subject Sample	# 1	0.136	01:40
Breath Volume		3.039	01:40
Air Blank		0.000	01:41
Air Blank		0.000	01:43
Subject Sample	# 2	0.136	01:43
Breath Volume		2.195	01:43
Air Blank		0.000	01:44
Control Test		0.078	01:44
Air Blank		0.000	01:45
Diagnostics Check		OK	01:45

Data Download Date/Time: 01/28/2020 13:22

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

INSTRUMENT REGISTERED TO: INDIALANTIC P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,ANTHONY,M
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: BLAZEWICZ,A
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 12/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:39

SUBJECT'S LAST NAME: WINCHESTER

DATE/TIME OF TEST: 12/21/2019 22:35:24

Diagnostics Check		OK	22:29
Air Blank		0.000	22:30
Control Test		0.080	22:30
Air Blank		0.000	22:30
Subject Sample	# 1	0.000	22:31
Breath Volume		2.664	22:31
Air Blank		0.000	22:31
Air Blank		0.000	22:33
Subject Sample	# 2	0.000	22:34
Breath Volume		3.359	22:34
Air Blank		0.000	22:34
Control Test		0.078	22:34
Air Blank		0.000	22:35
Diagnostics Check		OK	22:35

Data Download Date/Time: 01/28/2020 13:22

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKSON,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: JACKSON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 11/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: LOPEZ-BENAVENTE

DATE/TIME OF TEST: 12/01/2019 04:18:34

Diagnostics Check		OK	04:12
Air Blank		0.000	04:12
Control Test		0.082	04:13
Air Blank		0.000	04:13
Subject Sample	# 1	0.010	04:14
Breath Volume		1.718	04:14
Air Blank		0.000	04:14
Air Blank		0.000	04:16
Subject Sample	# 2	0.007	04:17
Breath Volume		1.175	04:17
Air Blank		0.000	04:17
Control Test		0.081	04:17
Air Blank		0.000	04:18
Diagnostics Check		OK	04:18

Data Download Date/Time: 12/06/2019 12:10

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,DWIGHT,A
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: WILSON,KALEB
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 11/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: RAMIREZ-NAZARIO

DATE/TIME OF TEST: 12/08/2019 22:57:53

Diagnostics Check		OK	22:51
Air Blank		0.000	22:52
Control Test		0.080	22:52
Air Blank		0.000	22:52
Subject Sample	# 1	0.198	22:53
Breath Volume		3.167	22:53
Air Blank		0.000	22:54
Air Blank		0.000	22:55
Subject Sample	# 2	0.183	22:56
Breath Volume		1.582	22:56
Air Blank		0.000	22:56
Control Test		0.082	22:57
Air Blank		0.000	22:57
Diagnostics Check		OK	22:57

Data Download Date/Time: 01/08/2020 08:16

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKSON,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: JACKSON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 11/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: COURT ORD
ARREST TIME: 08:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:45

SUBJECT'S LAST NAME: KOEHLER

DATE/TIME OF TEST: 12/12/2019 09:20:03

Diagnostics Check		OK	09:14
Air Blank		0.000	09:14
Control Test		0.081	09:15
Air Blank		0.000	09:15
Subject Sample	# 1	0.000	09:15
Breath Volume		2.761	09:15
Air Blank		0.000	09:16
Air Blank		0.000	09:18
Subject Sample	# 2	0.000	09:18
Breath Volume		2.789	09:18
Air Blank		0.000	09:19
Control Test		0.082	09:19
Air Blank		0.000	09:19
Diagnostics Check		OK	09:20

Data Download Date/Time: 01/08/2020 08:16

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKSON,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: JACKSON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 11/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: COURT ORD
ARREST TIME: 10:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:00

SUBJECT'S LAST NAME: ROMAN ROSA

DATE/TIME OF TEST: 12/12/2019 10:48:54

Diagnostics Check		OK	10:43
Air Blank		0.000	10:43
Control Test		0.082	10:43
Air Blank		0.000	10:44
Subject Sample	# 1	0.000	10:44
Breath Volume		2.695	10:44
Air Blank		0.000	10:45
Air Blank		0.000	10:47
Subject Sample	# 2	0.000	10:47
Breath Volume		2.148	10:47
Air Blank		0.000	10:47
Control Test		0.083	10:48
Air Blank		0.000	10:48
Diagnostics Check		OK	10:48

Data Download Date/Time: 01/08/2020 08:16

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKSON,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: JACKSON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 11/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: COURT ORD
ARREST TIME: 10:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:30

SUBJECT'S LAST NAME: DEEMER

DATE/TIME OF TEST: 12/12/2019 11:16:01

Diagnostics Check		OK	11:09
Air Blank		0.000	11:10
Control Test		0.082	11:10
Air Blank		0.000	11:10
Subject Sample	# 1	0.000	11:11
Breath Volume		1.472	11:11
Air Blank		0.000	11:11
Air Blank		0.000	11:13
Subject Sample	# 2	0.000	11:14
Breath Volume		1.945	11:14
Air Blank		0.000	11:15
Control Test		0.082	11:15
Air Blank		0.000	11:15
Diagnostics Check		OK	11:16

Data Download Date/Time: 01/08/2020 08:16

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARING,KELLY,A
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: WARING,KELLY
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 11/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: PROB/PAROL
ARREST TIME: 13:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:02

SUBJECT'S LAST NAME: MOLINA ORSINI

DATE/TIME OF TEST: 12/12/2019 14:25:54

Diagnostics Check		OK	14:19
Air Blank		0.000	14:20
Control Test		0.082	14:20
Air Blank		0.000	14:21
Subject Sample	# 1	0.000	14:21
Breath Volume		1.847	14:21
Air Blank		0.000	14:22
Air Blank		0.000	14:24
Subject Sample	# 2	0.000	14:24
Breath Volume		2.472	14:24
Air Blank		0.000	14:24
Control Test		0.084	14:25
Air Blank		0.000	14:25
Diagnostics Check		OK	14:25

Data Download Date/Time: 01/08/2020 08:16

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: BRUNER,SEAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 11/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: ENCARNACION

DATE/TIME OF TEST: 12/13/2019 23:23:26

Diagnostics Check		OK	23:10
Air Blank		0.000	23:11
Control Test		0.082	23:11
Air Blank		0.000	23:11
Subject Sample	# 1	0.200	23:14
Breath Volume		1.117	23:14
Air Blank		0.000	23:14
Air Blank		0.000	23:16
Subject Sample	# 2	0.222	23:17
Breath Volume		1.156	23:17
Air Blank		0.000	23:18
Air Blank		0.000	23:19
Subject Sample	# 3	0.193	23:21
Breath Volume		1.257	23:21
Air Blank		0.000	23:22
Control Test		0.080	23:22
Air Blank		0.000	23:23
Diagnostics Check		OK	23:23

Data Download Date/Time: 01/08/2020 08:16

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: BRUNER,SEAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 11/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 05:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: MANZARI

DATE/TIME OF TEST: 12/14/2019 05:45:22

Diagnostics Check		OK	05:37
Air Blank		0.000	05:38
Control Test		0.081	05:38
Air Blank		0.000	05:38
Subject Sample	# 1	0.129	05:40
Breath Volume		1.628	05:40
Air Blank		0.000	05:41
Air Blank		0.000	05:42
Subject Sample	# 2	0.131	05:43
Breath Volume		1.773	05:43
Air Blank		0.000	05:44
Control Test		0.083	05:44
Air Blank		0.000	05:45
Diagnostics Check		OK	05:45

Data Download Date/Time: 01/08/2020 08:16

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: BRUNER,SEAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 11/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 12/15/2019 03:38:56

Diagnostics Check		OK	03:32
Air Blank		0.000	03:32
Control Test		0.082	03:33
Air Blank		0.000	03:33
Subject Sample	# 1	0.102	03:34
Breath Volume		1.203	03:34
Air Blank		0.000	03:34
Air Blank		0.000	03:36
Subject Sample	# 2	0.108	03:37
Breath Volume		1.886	03:37
Air Blank		0.000	03:38
Control Test		0.082	03:38
Air Blank		0.000	03:38
Diagnostics Check		OK	03:38

Data Download Date/Time: 01/08/2020 08:16

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: BRUNER,SEAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 11/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: RAMKISHUN

DATE/TIME OF TEST: 12/16/2019 02:49:37

Diagnostics Check		OK	02:42
Air Blank		0.000	02:42
Control Test		0.081	02:43
Air Blank		0.000	02:43
Subject Sample	# 1	0.199	02:44
Breath Volume		1.574	02:44
Air Blank		0.000	02:44
Air Blank		0.000	02:46
Subject Sample	# 2	0.204	02:48
Breath Volume		2.238	02:48
Air Blank		0.000	02:48
Control Test		0.079	02:49
Air Blank		0.000	02:49
Diagnostics Check		OK	02:49

Data Download Date/Time: 01/08/2020 08:16

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: BRUNER,SEAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 11/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 19:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:27

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 12/19/2019 19:57:30

Diagnostics Check		OK	19:51
Air Blank		0.000	19:51
Control Test		0.082	19:52
Air Blank		0.000	19:52
Subject Sample	# 1	0.218	19:53
Breath Volume		1.640	19:53
Air Blank		0.000	19:53
Air Blank		0.000	19:55
Subject Sample	# 2	0.223	19:55
Breath Volume		1.828	19:55
Air Blank		0.000	19:56
Control Test		0.080	19:56
Air Blank		0.000	19:57
Diagnostics Check		OK	19:57

Data Download Date/Time: 01/08/2020 08:16

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,LATORIAN,T
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: TURNER,LATORIAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 11/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:06

SUBJECT'S LAST NAME: MALDONADO RIVERA

DATE/TIME OF TEST: 12/25/2019 22:41:15

Diagnostics Check		OK	22:35
Air Blank		0.000	22:35
Control Test		0.081	22:36
Air Blank		0.000	22:36
Subject Sample	# 1	0.201	22:36
Breath Volume		1.441	22:36
Air Blank		0.000	22:37
Air Blank		0.000	22:39
Subject Sample	# 2	0.187	22:39
Breath Volume		1.851	22:39
Air Blank		0.000	22:40
Control Test		0.078	22:40
Air Blank		0.000	22:41
Diagnostics Check		OK	22:41

Data Download Date/Time: 01/08/2020 08:16

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,STEVEN,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: EVANS,STEVEN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 11/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: MUNIZ SANTIAGO

DATE/TIME OF TEST: 12/26/2019 03:57:11

Diagnostics Check		OK	03:42	
Air Blank		0.000	03:42	
Control Test		0.079	03:42	
Air Blank		0.000	03:43	
Subject Sample	# 1	0.236	03:44	
Breath Volume		1.660	03:44	
Air Blank		0.000	03:44	
Air Blank		0.000	03:46	
Subject Sample	# 2	0.206	03:49	Volume Not Met
Breath Volume		0.949	03:49	
Air Blank		0.000	03:50	
Air Blank		0.000	03:52	
Subject Sample	# 3	0.205	03:55	Volume Not Met
Breath Volume		0.734	03:55	
Air Blank		0.000	03:56	
Control Test		0.081	03:56	
Air Blank		0.000	03:57	
Diagnostics Check		OK	03:57	

Data Download Date/Time: 01/08/2020 08:16

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,STEVEN,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: EVANS,STEVEN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 11/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: MUNIZ SANTIAGO

DATE/TIME OF TEST: 12/26/2019 04:08:27

Diagnostics Check		OK	04:03	
Air Blank		0.000	04:04	
Control Test		0.080	04:04	
Air Blank		0.000	04:05	
Subject Sample	# 1		04:06	Subject Test Refused
Breath Volume		0.875	04:06	
Air Blank		0.000	04:07	
Control Test		0.081	04:07	
Air Blank		0.000	04:08	
Diagnostics Check		OK	04:08	

Data Download Date/Time: 01/08/2020 08:16

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,STEVEN,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: EVANS,STEVEN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 11/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: MUNIZ SANTIAGO

DATE/TIME OF TEST: 12/26/2019 04:30:35

Diagnostics Check		OK	04:22
Air Blank		0.000	04:22
Control Test		0.078	04:23
Air Blank		0.000	04:23
Subject Sample	# 1	0.214	04:24
Breath Volume		1.816	04:24
Air Blank		0.000	04:25
Air Blank		0.000	04:26
Subject Sample	# 2	0.212	04:29
Breath Volume		1.214	04:29
Air Blank		0.000	04:29
Control Test		0.080	04:29
Air Blank		0.000	04:30
Diagnostics Check		OK	04:30

Data Download Date/Time: 01/08/2020 08:17

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORJAS,NOE,A
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: BERRY,VICTORIA
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 11/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 12/06/2019 21:51:25

Diagnostics Check		OK	21:48	
Air Blank		0.000	21:48	
Control Test		0.080	21:48	
Air Blank		0.000	21:49	
Subject Sample	# 1	0.000	21:49	
Breath Volume		4.027	21:49	
Air Blank			21:50	Improper Sample
Air Blank		0.000	21:50	
Control Test		0.080	21:50	
Air Blank		0.000	21:51	
Diagnostics Check		OK	21:51	

Data Download Date/Time: 12/30/2019 13:15

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORJAS,NOE,A
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: BERRY,VICTORIA
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 11/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 12/06/2019 22:02:54

Diagnostics Check		OK	21:54
Air Blank		0.000	21:55
Control Test		0.079	21:55
Air Blank		0.000	21:56
Subject Sample	# 1	0.000	21:58
Breath Volume		3.117	21:58
Air Blank		0.000	21:59
Air Blank		0.000	22:01
Subject Sample	# 2	0.000	22:01
Breath Volume		2.230	22:01
Air Blank		0.000	22:01
Control Test		0.079	22:02
Air Blank		0.000	22:02
Diagnostics Check		OK	22:02

Data Download Date/Time: 12/30/2019 13:15

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORJAS,NOE,A
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: SAMUELS,ANDREW
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 11/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: SCHAFER

DATE/TIME OF TEST: 12/07/2019 00:59:20

Diagnostics Check		OK	00:50
Air Blank		0.000	00:50
Control Test		0.079	00:51
Air Blank		0.000	00:51
Subject Sample	# 1	0.313	00:51
Breath Volume		2.437	00:51
Air Blank		0.000	00:52
Air Blank		0.000	00:54
Subject Sample	# 2	0.289	00:54
Breath Volume		2.808	00:54
Air Blank		0.000	00:55
Air Blank		0.000	00:57
Subject Sample	# 3	0.288	00:57
Breath Volume		3.421	00:57
Air Blank		0.000	00:58
Control Test		0.078	00:58
Air Blank		0.000	00:59
Diagnostics Check		OK	00:59

Data Download Date/Time: 12/30/2019 13:15

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORJAS,NOE,A
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: BERRY,VICTORIA
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 11/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:04

SUBJECT'S LAST NAME: RIKER

DATE/TIME OF TEST: 12/07/2019 21:36:42

Diagnostics Check		OK	21:30
Air Blank		0.000	21:30
Control Test		0.079	21:31
Air Blank		0.000	21:31
Subject Sample	# 1	0.141	21:32
Breath Volume		3.792	21:32
Air Blank		0.000	21:32
Air Blank		0.000	21:34
Subject Sample	# 2	0.143	21:35
Breath Volume		4.527	21:35
Air Blank		0.000	21:35
Control Test		0.079	21:36
Air Blank		0.000	21:36
Diagnostics Check		OK	21:36

Data Download Date/Time: 12/30/2019 13:15

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOGAN,JONATHAN,
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: MCDONALD,COLE
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 11/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: WARD

DATE/TIME OF TEST: 12/10/2019 02:56:14

Diagnostics Check		OK	02:50
Air Blank		0.000	02:50
Control Test		0.080	02:50
Air Blank		0.000	02:51
Subject Sample	# 1	0.011	02:51
Breath Volume		1.886	02:51
Air Blank		0.000	02:52
Air Blank		0.000	02:54
Subject Sample	# 2	0.009	02:54
Breath Volume		1.675	02:54
Air Blank		0.000	02:55
Control Test		0.079	02:55
Air Blank		0.000	02:56
Diagnostics Check		OK	02:56

Data Download Date/Time: 12/30/2019 13:15

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOGAN,JONATHAN,
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: LOGAN,JONATHAN
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 11/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: FICK

DATE/TIME OF TEST: 12/15/2019 00:48:01

Diagnostics Check		OK	00:41
Air Blank		0.000	00:41
Control Test		0.079	00:41
Air Blank		0.000	00:42
Subject Sample	# 1	0.122	00:42
Breath Volume		1.355	00:42
Air Blank		0.000	00:43
Air Blank		0.000	00:45
Subject Sample	# 2	0.123	00:46
Breath Volume		1.550	00:46
Air Blank		0.000	00:47
Control Test		0.079	00:47
Air Blank		0.000	00:47
Diagnostics Check		OK	00:48

Data Download Date/Time: 12/30/2019 13:15

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOGAN,JONATHAN,
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: SAMUELS,ANDREW
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 11/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 12/21/2019 04:40:17

Diagnostics Check		OK	04:34
Air Blank		0.000	04:34
Control Test		0.079	04:34
Air Blank		0.000	04:35
Subject Sample	# 1	0.138	04:35
Breath Volume		1.738	04:35
Air Blank		0.000	04:36
Air Blank		0.000	04:38
Subject Sample	# 2	0.134	04:38
Breath Volume		1.476	04:38
Air Blank		0.000	04:39
Control Test		0.079	04:39
Air Blank		0.000	04:40
Diagnostics Check		OK	04:40

Data Download Date/Time: 12/30/2019 13:15

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOGAN,JONATHAN,
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: SAMUELS,ANDREW
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 12/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:39

SUBJECT'S LAST NAME: MANGUS

DATE/TIME OF TEST: 12/30/2019 20:19:54

Diagnostics Check		OK	20:13
Air Blank		0.000	20:14
Control Test		0.080	20:14
Air Blank		0.000	20:14
Subject Sample	# 1	0.315	20:15
Breath Volume		1.808	20:15
Air Blank		0.000	20:16
Air Blank		0.000	20:17
Subject Sample	# 2	0.317	20:18
Breath Volume		1.660	20:18
Air Blank		0.000	20:18
Control Test		0.078	20:19
Air Blank		0.000	20:19
Diagnostics Check		OK	20:19

Data Download Date/Time: 01/28/2020 17:13

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: GARCIA PEREZ

DATE/TIME OF TEST: 12/01/2019 03:39:56

Diagnostics Check		OK	03:32
Air Blank		0.000	03:33
Control Test		0.079	03:33
Air Blank		0.000	03:34
Subject Sample	# 1	0.097	03:35
Breath Volume		1.554	03:35
Air Blank		0.000	03:35
Air Blank		0.000	03:37
Subject Sample	# 2	0.097	03:38
Breath Volume		1.335	03:38
Air Blank		0.000	03:38
Control Test		0.079	03:39
Air Blank		0.000	03:39
Diagnostics Check		OK	03:39

Data Download Date/Time: 12/03/2019 10:51

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEPHERD,P,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VO,S
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 09:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:54

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 12/06/2019 11:24:05

Diagnostics Check		OK	11:17
Air Blank		0.000	11:18
Control Test		0.080	11:18
Air Blank		0.000	11:19
Subject Sample	# 1	0.000	11:19
Breath Volume		2.113	11:19
Air Blank		0.000	11:20
Air Blank		0.000	11:22
Subject Sample	# 2	0.000	11:22
Breath Volume		1.832	11:22
Air Blank		0.000	11:23
Control Test		0.079	11:23
Air Blank		0.000	11:23
Diagnostics Check		OK	11:24

Data Download Date/Time: 12/20/2019 10:25

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 12/10/2019 02:58:30

Diagnostics Check		OK	02:52
Air Blank		0.000	02:52
Control Test		0.079	02:53
Air Blank		0.000	02:53
Subject Sample	# 1	0.000	02:54
Breath Volume		2.433	02:54
Air Blank		0.000	02:54
Air Blank		0.000	02:56
Subject Sample	# 2	0.000	02:57
Breath Volume		4.042	02:57
Air Blank		0.000	02:57
Control Test		0.079	02:57
Air Blank		0.000	02:58
Diagnostics Check		OK	02:58

Data Download Date/Time: 12/20/2019 10:25

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DASHER,JOSHUA,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: KMOWLES

DATE/TIME OF TEST: 12/10/2019 22:24:24

Diagnostics Check		OK	22:17
Air Blank		0.000	22:18
Control Test		0.079	22:18
Air Blank		0.000	22:19
Subject Sample	# 1	0.032	22:19
Breath Volume		1.847	22:19
Air Blank		0.000	22:20
Air Blank		0.000	22:22
Subject Sample	# 2	0.029	22:22
Breath Volume		2.046	22:22
Air Blank		0.000	22:23
Control Test		0.079	22:23
Air Blank		0.000	22:24
Diagnostics Check		OK	22:24

Data Download Date/Time: 12/20/2019 10:25

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DASHER,JOSHUA,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNS,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: MALCOLM

DATE/TIME OF TEST: 12/11/2019 01:42:28

Diagnostics Check		OK	01:36
Air Blank		0.000	01:36
Control Test		0.079	01:37
Air Blank		0.000	01:37
Subject Sample	# 1	0.000	01:38
Breath Volume		3.187	01:38
Air Blank		0.000	01:38
Air Blank		0.000	01:40
Subject Sample	# 2	0.000	01:41
Breath Volume		2.253	01:41
Air Blank		0.000	01:41
Control Test		0.079	01:41
Air Blank		0.000	01:42
Diagnostics Check		OK	01:42

Data Download Date/Time: 12/20/2019 10:25

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 12/13/2019 01:34:31

Diagnostics Check		OK	01:31	
Air Blank		0.000	01:32	
Control Test		0.079	01:32	
Air Blank		0.000	01:32	
Subject Sample	# 1		01:33	Subject Test Refused
Breath Volume		0.000	01:33	
Air Blank		0.000	01:33	
Control Test		0.079	01:33	
Air Blank		0.000	01:34	
Diagnostics Check		OK	01:34	

Data Download Date/Time: 12/20/2019 10:25

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SOSA,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:47

SUBJECT'S LAST NAME: STONE

DATE/TIME OF TEST: 12/13/2019 22:14:32

Diagnostics Check		OK	22:07
Air Blank		0.000	22:08
Control Test		0.078	22:08
Air Blank		0.000	22:09
Subject Sample	# 1	0.215	22:09
Breath Volume		1.808	22:09
Air Blank		0.000	22:10
Air Blank		0.000	22:12
Subject Sample	# 2	0.207	22:12
Breath Volume		1.648	22:12
Air Blank		0.000	22:13
Control Test		0.078	22:13
Air Blank		0.000	22:14
Diagnostics Check		OK	22:14

Data Download Date/Time: 12/20/2019 10:25

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEPHERD,P,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MARQUEZ,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:50

SUBJECT'S LAST NAME: NICHOLS

DATE/TIME OF TEST: 12/14/2019 07:24:38

Diagnostics Check		OK	07:12	
Air Blank		0.000	07:12	
Control Test		0.078	07:13	
Air Blank		0.000	07:13	
Subject Sample	# 1		07:16	No Sample Provided
Breath Volume		0.000	07:16	
Air Blank		0.000	07:17	
Air Blank		0.000	07:19	
Subject Sample	# 2	0.120	07:19	No .020 agreement
Breath Volume		3.957	07:19	
Air Blank		0.000	07:20	
Air Blank		0.000	07:22	
Subject Sample	# 3	0.089	07:22	No .020 agreement
Breath Volume		1.304	07:22	
Air Blank		0.000	07:23	
Control Test		0.078	07:23	
Air Blank		0.000	07:24	
Diagnostics Check		OK	07:24	

Data Download Date/Time: 12/20/2019 10:25

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEPHERD,P,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MARQUEZ,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:50

SUBJECT'S LAST NAME: NICHOLS

DATE/TIME OF TEST: 12/14/2019 07:32:08

Diagnostics Check		OK	07:29	
Air Blank		0.000	07:29	
Control Test		0.078	07:29	
Air Blank		0.000	07:30	
Subject Sample	# 1		07:30	Subject Test Refused
Breath Volume		0.000	07:30	
Air Blank		0.000	07:31	
Control Test		0.077	07:31	
Air Blank		0.000	07:32	
Diagnostics Check		OK	07:32	

Data Download Date/Time: 12/20/2019 10:25

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOINER,D
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 12/15/2019 01:24:29

Diagnostics Check		OK	01:17
Air Blank		0.000	01:18
Control Test		0.079	01:18
Air Blank		0.000	01:18
Subject Sample	# 1	0.171	01:19
Breath Volume		2.070	01:19
Air Blank		0.000	01:20
Air Blank		0.000	01:22
Subject Sample	# 2	0.168	01:22
Breath Volume		2.152	01:22
Air Blank		0.000	01:23
Control Test		0.078	01:23
Air Blank		0.000	01:24
Diagnostics Check		OK	01:24

Data Download Date/Time: 12/20/2019 10:25

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: REEVES

DATE/TIME OF TEST: 12/15/2019 23:55:10

Diagnostics Check		OK	23:52	
Air Blank		0.000	23:52	
Control Test		0.079	23:53	
Air Blank		0.000	23:53	
Subject Sample	# 1		23:53	Subject Test Refused
Breath Volume		0.000	23:53	
Air Blank		0.000	23:54	
Control Test		0.079	23:54	
Air Blank		0.000	23:55	
Diagnostics Check		OK	23:55	

Data Download Date/Time: 12/20/2019 10:25

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DZAMKO,J
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: FIELDS

DATE/TIME OF TEST: 12/16/2019 04:05:06

Diagnostics Check		OK	03:58
Air Blank		0.000	03:58
Control Test		0.080	03:59
Air Blank		0.000	03:59
Subject Sample	# 1	0.106	04:00
Breath Volume		2.191	04:00
Air Blank		0.000	04:01
Air Blank		0.000	04:02
Subject Sample	# 2	0.105	04:03
Breath Volume		2.632	04:03
Air Blank		0.000	04:04
Control Test		0.079	04:04
Air Blank		0.000	04:05
Diagnostics Check		OK	04:05

Data Download Date/Time: 12/20/2019 10:26

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SANDIFORD,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: BECK

DATE/TIME OF TEST: 12/16/2019 23:15:02

Diagnostics Check		OK	23:12	
Air Blank		0.000	23:12	
Control Test		0.078	23:12	
Air Blank		0.000	23:13	
Subject Sample	# 1		23:13	Subject Test Refused
Breath Volume		0.000	23:13	
Air Blank		0.000	23:14	
Control Test		0.079	23:14	
Air Blank		0.000	23:14	
Diagnostics Check		OK	23:15	

Data Download Date/Time: 12/20/2019 10:26

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 12/17/2019 01:59:49

Diagnostics Check		OK	01:52
Air Blank		0.000	01:53
Control Test		0.079	01:53
Air Blank		0.000	01:54
Subject Sample	# 1	0.311	01:55
Breath Volume		3.167	01:55
Air Blank		0.000	01:55
Air Blank		0.000	01:57
Subject Sample	# 2	0.320	01:58
Breath Volume		2.375	01:58
Air Blank		0.000	01:58
Control Test		0.077	01:59
Air Blank		0.000	01:59
Diagnostics Check		OK	01:59

Data Download Date/Time: 12/20/2019 10:26

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CLARK,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: MARONSKI

DATE/TIME OF TEST: 12/17/2019 05:18:10

Diagnostics Check		OK	05:15	
Air Blank		0.000	05:15	
Control Test		0.079	05:16	
Air Blank		0.000	05:16	
Subject Sample	# 1		05:16	Subject Test Refused
Breath Volume		0.000	05:16	
Air Blank		0.000	05:17	
Control Test		0.079	05:17	
Air Blank		0.000	05:18	
Diagnostics Check		OK	05:18	

Data Download Date/Time: 12/20/2019 10:26

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAY,L,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:27

SUBJECT'S LAST NAME: CLAUDIOVAZQUEZ

DATE/TIME OF TEST: 12/18/2019 18:55:14

Diagnostics Check		OK	18:48
Air Blank		0.000	18:48
Control Test		0.079	18:49
Air Blank		0.000	18:49
Subject Sample	# 1	0.103	18:50
Breath Volume		1.335	18:50
Air Blank		0.000	18:51
Air Blank		0.000	18:53
Subject Sample	# 2	0.108	18:53
Breath Volume		1.214	18:53
Air Blank		0.000	18:54
Control Test		0.079	18:54
Air Blank		0.000	18:55
Diagnostics Check		OK	18:55

Data Download Date/Time: 12/20/2019 10:26

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRUETT,RICHARD,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SCHERSCHEL,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 06:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:43

SUBJECT'S LAST NAME: HADDAN

DATE/TIME OF TEST: 12/21/2019 08:08:04

Diagnostics Check		OK	08:04	
Air Blank		0.000	08:04	
Control Test		0.078	08:05	
Air Blank		0.000	08:05	
Subject Sample	# 1		08:06	Subject Test Refused
Breath Volume		0.000	08:06	
Air Blank		0.000	08:07	
Control Test		0.077	08:07	
Air Blank		0.000	08:07	
Diagnostics Check		OK	08:08	

Data Download Date/Time: 01/16/2020 07:56

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: FEWOX

DATE/TIME OF TEST: 12/25/2019 02:48:56

Diagnostics Check		OK	02:41
Air Blank		0.000	02:42
Control Test		0.080	02:42
Air Blank		0.000	02:43
Subject Sample	# 1	0.211	02:43
Breath Volume		4.207	02:43
Air Blank		0.000	02:44
Air Blank		0.000	02:46
Subject Sample	# 2	0.208	02:47
Breath Volume		3.933	02:47
Air Blank		0.000	02:47
Control Test		0.078	02:48
Air Blank		0.000	02:48
Diagnostics Check		OK	02:48

Data Download Date/Time: 01/16/2020 07:56

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: PENA

DATE/TIME OF TEST: 12/26/2019 22:54:38

Diagnostics Check		OK	22:51	
Air Blank		0.000	22:52	
Control Test		0.079	22:52	
Air Blank		0.000	22:52	
Subject Sample	# 1		22:53	Subject Test Refused
Breath Volume		0.000	22:53	
Air Blank		0.000	22:53	
Control Test		0.079	22:54	
Air Blank		0.000	22:54	
Diagnostics Check		OK	22:54	

Data Download Date/Time: 01/16/2020 07:56

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAY,L,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ALEXANDER,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 11:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:03

SUBJECT'S LAST NAME: MAYHEW

DATE/TIME OF TEST: 12/27/2019 13:36:21

Diagnostics Check		OK	13:26
Air Blank		0.000	13:27
Control Test		0.078	13:27
Air Blank		0.000	13:28
Subject Sample	# 1	0.239	13:30
Breath Volume		2.203	13:30
Air Blank		0.000	13:31
Air Blank		0.000	13:32
Subject Sample	# 2	0.222	13:34
Breath Volume		1.484	13:34
Air Blank		0.000	13:35
Control Test		0.077	13:35
Air Blank		0.000	13:36
Diagnostics Check		OK	13:36

Data Download Date/Time: 01/16/2020 07:56

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAY,L,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TUCKER,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 13:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:21

SUBJECT'S LAST NAME: MILLINGTON

DATE/TIME OF TEST: 12/30/2019 15:48:44

Diagnostics Check		OK	15:42
Air Blank		0.000	15:42
Control Test		0.079	15:43
Air Blank		0.000	15:43
Subject Sample	# 1	0.000	15:44
Breath Volume		1.625	15:44
Air Blank		0.000	15:44
Air Blank		0.000	15:46
Subject Sample	# 2	0.000	15:47
Breath Volume		1.398	15:47
Air Blank		0.000	15:47
Control Test		0.079	15:48
Air Blank		0.000	15:48
Diagnostics Check		OK	15:48

Data Download Date/Time: 01/16/2020 07:56

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAY,L,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VO,S
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 15:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:01

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 12/30/2019 17:28:35

Diagnostics Check		OK	17:21
Air Blank		0.000	17:22
Control Test		0.079	17:22
Air Blank		0.000	17:22
Subject Sample	# 1	0.203	17:23
Breath Volume		1.292	17:23
Air Blank		0.000	17:24
Air Blank		0.000	17:26
Subject Sample	# 2	0.206	17:26
Breath Volume		1.546	17:26
Air Blank		0.000	17:27
Control Test		0.078	17:27
Air Blank		0.000	17:28
Diagnostics Check		OK	17:28

Data Download Date/Time: 01/16/2020 07:56

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: MARTIN MAYE

DATE/TIME OF TEST: 12/06/2019 23:59:35

Diagnostics Check		OK	23:52
Air Blank		0.000	23:53
Control Test		0.079	23:53
Air Blank		0.000	23:54
Subject Sample	# 1	0.113	23:55
Breath Volume		2.355	23:55
Air Blank		0.000	23:55
Air Blank		0.000	23:57
Subject Sample	# 2	0.111	23:58
Breath Volume		2.214	23:58
Air Blank		0.000	23:58
Control Test		0.080	23:59
Air Blank		0.000	23:59
Diagnostics Check		OK	23:59

Data Download Date/Time: 01/10/2020 11:20

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: LAWRENCE

DATE/TIME OF TEST: 12/07/2019 01:13:04

Diagnostics Check		OK	01:03
Air Blank		0.000	01:04
Control Test		0.079	01:04
Air Blank		0.000	01:04
Subject Sample	# 1	0.152	01:08
Breath Volume		1.937	01:08
Air Blank		0.000	01:08
Air Blank		0.000	01:10
Subject Sample	# 2	0.139	01:11
Breath Volume		1.128	01:11
Air Blank		0.000	01:12
Control Test		0.080	01:12
Air Blank		0.000	01:12
Diagnostics Check		OK	01:13

Data Download Date/Time: 01/10/2020 11:20

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUFF,ZACHARY,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: HUFF,ZACHARY
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: STACK

DATE/TIME OF TEST: 12/07/2019 01:43:57

Diagnostics Check		OK	01:40	
Air Blank		0.000	01:41	
Control Test		0.079	01:41	
Air Blank		0.000	01:42	
Subject Sample	# 1		01:42	Subject Test Refused
Breath Volume		0.000	01:42	
Air Blank		0.000	01:43	
Control Test		0.079	01:43	
Air Blank		0.000	01:43	
Diagnostics Check		OK	01:43	

Data Download Date/Time: 01/10/2020 11:20

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 12/08/2019 03:48:03

Diagnostics Check		OK	03:45	
Air Blank		0.000	03:45	
Control Test		0.079	03:46	
Air Blank		0.000	03:46	
Subject Sample	# 1		03:46	Subject Test Refused
Breath Volume		0.000	03:46	
Air Blank		0.000	03:47	
Control Test		0.079	03:47	
Air Blank		0.000	03:47	
Diagnostics Check		OK	03:48	

Data Download Date/Time: 01/10/2020 11:20

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUFF,ZACHARY,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: HUFF,ZACHARY
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: FLORENCE

DATE/TIME OF TEST: 12/08/2019 23:33:14

Diagnostics Check		OK	23:21	
Air Blank		0.000	23:21	
Control Test		0.079	23:21	
Air Blank		0.000	23:22	
Subject Sample	# 1	0.113	23:25	Volume Not Met
Breath Volume		0.683	23:25	
Air Blank		0.000	23:26	
Air Blank		0.000	23:27	
Subject Sample	# 2	0.107	23:28	
Breath Volume		1.226	23:28	
Air Blank		0.000	23:29	
Air Blank		0.000	23:31	
Subject Sample	# 3	0.110	23:31	
Breath Volume		1.300	23:31	
Air Blank		0.000	23:32	
Control Test		0.079	23:32	
Air Blank		0.000	23:33	
Diagnostics Check		OK	23:33	

Data Download Date/Time: 01/10/2020 11:20

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCORDINO,DOUGLAS,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SCORDINO,DOUGLAS
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: BOYT

DATE/TIME OF TEST: 12/10/2019 03:19:59

Diagnostics Check		OK	03:11	
Air Blank		0.000	03:11	
Control Test		0.079	03:11	
Air Blank		0.000	03:12	
Subject Sample	# 1		03:15	No Sample Provided
Breath Volume		0.000	03:15	
Air Blank		0.000	03:15	
Air Blank		0.000	03:17	
Subject Sample	# 2		03:18	Subject Test Refused
Breath Volume		0.000	03:18	
Air Blank		0.000	03:19	
Control Test		0.078	03:19	
Air Blank		0.000	03:19	
Diagnostics Check		OK	03:19	

Data Download Date/Time: 01/10/2020 11:20

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOEDE,BRANDON,W
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: DIAZ,A
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: BOZEMAN

DATE/TIME OF TEST: 12/10/2019 23:21:24

Diagnostics Check		OK	23:18	
Air Blank		0.000	23:18	
Control Test		0.080	23:19	
Air Blank		0.000	23:19	
Subject Sample	# 1		23:20	Subject Test Refused
Breath Volume		0.000	23:20	
Air Blank		0.000	23:20	
Control Test		0.079	23:20	
Air Blank		0.000	23:21	
Diagnostics Check		OK	23:21	

Data Download Date/Time: 01/10/2020 11:20

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: GRANT

DATE/TIME OF TEST: 12/12/2019 00:30:27

Diagnostics Check		OK	00:22
Air Blank		0.000	00:23
Control Test		0.080	00:23
Air Blank		0.000	00:24
Subject Sample	# 1	0.117	00:26
Breath Volume		2.390	00:26
Air Blank		0.000	00:26
Air Blank		0.000	00:28
Subject Sample	# 2	0.113	00:29
Breath Volume		1.394	00:29
Air Blank		0.000	00:29
Control Test		0.079	00:29
Air Blank		0.000	00:30
Diagnostics Check		OK	00:30

Data Download Date/Time: 01/10/2020 11:20

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 12/12/2019 02:01:34

Diagnostics Check		OK	01:53	
Air Blank		0.000	01:53	
Control Test		0.079	01:54	
Air Blank		0.000	01:54	
Subject Sample	# 1	0.248	01:55	
Breath Volume		1.937	01:55	
Air Blank		0.000	01:56	
Air Blank		0.000	01:58	
Subject Sample	# 2		02:00	Subject Test Refused
Breath Volume		0.000	02:00	
Air Blank		0.000	02:00	
Control Test		0.078	02:00	
Air Blank		0.000	02:01	
Diagnostics Check		OK	02:01	

Data Download Date/Time: 01/10/2020 11:20

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: GUZMAN

DATE/TIME OF TEST: 12/13/2019 00:02:02

Diagnostics Check		OK	23:54
Air Blank		0.000	23:55
Control Test		0.079	23:55
Air Blank		0.000	23:55
Subject Sample	# 2	0.242	23:57
Breath Volume		3.468	23:57
Air Blank		0.000	23:58
Air Blank		0.000	23:59
Subject Sample	# 1	0.248	00:00
Breath Volume		3.847	00:00
Air Blank		0.000	00:01
Control Test		0.081	00:01
Air Blank		0.000	00:01
Diagnostics Check		OK	00:02

Data Download Date/Time: 01/10/2020 11:20

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: ZURLA

DATE/TIME OF TEST: 12/14/2019 18:47:15

Diagnostics Check		OK	18:40
Air Blank		0.000	18:41
Control Test		0.079	18:41
Air Blank		0.000	18:41
Subject Sample	# 1	0.222	18:42
Breath Volume		1.335	18:42
Air Blank		0.000	18:43
Air Blank		0.000	18:45
Subject Sample	# 2	0.227	18:45
Breath Volume		1.238	18:45
Air Blank		0.000	18:46
Control Test		0.079	18:46
Air Blank		0.000	18:47
Diagnostics Check		OK	18:47

Data Download Date/Time: 01/10/2020 11:20

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: CHELF

DATE/TIME OF TEST: 12/14/2019 22:37:43

Diagnostics Check		OK	22:34	
Air Blank		0.000	22:35	
Control Test		0.078	22:35	
Air Blank		0.000	22:36	
Subject Sample	# 1		22:36	Subject Test Refused
Breath Volume		0.000	22:36	
Air Blank		0.000	22:36	
Control Test		0.079	22:37	
Air Blank		0.000	22:37	
Diagnostics Check		OK	22:37	

Data Download Date/Time: 01/10/2020 11:20

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINISH,BARRY,A
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: LANIER,TRISTEN
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: SHIREY

DATE/TIME OF TEST: 12/29/2019 02:39:52

Diagnostics Check		OK	02:33
Air Blank		0.000	02:34
Control Test		0.079	02:34
Air Blank		0.000	02:35
Subject Sample	# 1	0.255	02:35
Breath Volume		1.707	02:35
Air Blank		0.000	02:36
Air Blank		0.000	02:37
Subject Sample	# 2	0.255	02:38
Breath Volume		1.546	02:38
Air Blank		0.000	02:38
Control Test		0.079	02:39
Air Blank		0.000	02:39
Diagnostics Check		OK	02:39

Data Download Date/Time: 12/30/2019 21:11

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: COCONUT CREEK PD
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: COCONUT CREEK PD
LAST AGENCY INSPECTION: 11/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1094981
DRY GAS EXPIRY DATE: 02/22/2022
VIOLATION CODE: DUI
ARREST TIME: 17:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:51

SUBJECT'S LAST NAME: BETTS

DATE/TIME OF TEST: 12/05/2019 18:28:37

Diagnostics Check		OK	18:22
Air Blank		0.000	18:22
Control Test		0.080	18:23
Air Blank		0.000	18:23
Subject Sample	# 1	0.378	18:24
Breath Volume		2.207	18:24
Air Blank		0.000	18:24
Air Blank		0.000	18:26
Subject Sample	# 2	0.373	18:27
Breath Volume		1.812	18:27
Air Blank		0.000	18:27
Control Test		0.080	18:28
Air Blank		0.000	18:28
Diagnostics Check		OK	18:28

Data Download Date/Time: 12/30/2019 15:44

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,GAVIN,O
BREATH TEST OPERATOR AGENCY: COCONUT CREEK PD
ARRESTING OFFICER: LARA,JENNIFER
ARRESTING OFFICER AGENCY: COCONUT CREEK PD
LAST AGENCY INSPECTION: 11/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1094981
DRY GAS EXPIRY DATE: 02/22/2022
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: BRAND

DATE/TIME OF TEST: 12/06/2019 04:06:29

Diagnostics Check		OK	04:03	
Air Blank		0.000	04:03	
Control Test		0.081	04:03	
Air Blank		0.000	04:04	
Subject Sample	# 1		04:05	Subject Test Refused
Breath Volume		0.000	04:05	
Air Blank		0.000	04:05	
Control Test		0.080	04:05	
Air Blank		0.000	04:06	
Diagnostics Check		OK	04:06	

Data Download Date/Time: 12/30/2019 15:44

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: COCONUT CREEK PD
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: COCONUT CREEK PD
LAST AGENCY INSPECTION: 11/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1094981
DRY GAS EXPIRY DATE: 02/22/2022
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: COLEY

DATE/TIME OF TEST: 12/11/2019 00:05:54

Diagnostics Check		OK	23:59
Air Blank		0.000	00:00
Control Test		0.081	00:00
Air Blank		0.000	00:00
Subject Sample	# 1	0.065	00:01
Breath Volume		1.894	00:01
Air Blank		0.000	00:02
Air Blank		0.000	00:04
Subject Sample	# 2	0.064	00:04
Breath Volume		1.570	00:04
Air Blank		0.000	00:05
Control Test		0.080	00:05
Air Blank		0.000	00:05
Diagnostics Check		OK	00:05

Data Download Date/Time: 12/30/2019 15:44

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: COCONUT CREEK PD
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: COCONUT CREEK PD
LAST AGENCY INSPECTION: 11/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1094981
DRY GAS EXPIRY DATE: 02/22/2022
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: OLIPHANT

DATE/TIME OF TEST: 12/12/2019 02:32:11

Diagnostics Check		OK	02:29	
Air Blank		0.000	02:29	
Control Test		0.080	02:30	
Air Blank		0.000	02:30	
Subject Sample	# 1		02:30	Subject Test Refused
Breath Volume		0.000	02:30	
Air Blank		0.000	02:31	
Control Test		0.081	02:31	
Air Blank		0.000	02:32	
Diagnostics Check		OK	02:32	

Data Download Date/Time: 12/30/2019 15:44

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,DONALD,P
BREATH TEST OPERATOR AGENCY: COCONUT CREEK PD
ARRESTING OFFICER: DAVIS,DONALD
ARRESTING OFFICER AGENCY: COCONUT CREEK PD
LAST AGENCY INSPECTION: 11/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1094981
DRY GAS EXPIRY DATE: 02/22/2022
VIOLATION CODE: DUI
ARREST TIME: 08:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:58

SUBJECT'S LAST NAME: BERMAN

DATE/TIME OF TEST: 12/24/2019 09:39:47

Diagnostics Check		OK	09:36	
Air Blank		0.000	09:37	
Control Test		0.081	09:37	
Air Blank		0.000	09:38	
Subject Sample	# 1		09:38	Subject Test Refused
Breath Volume		0.000	09:38	
Air Blank		0.000	09:38	
Control Test		0.080	09:39	
Air Blank		0.000	09:39	
Diagnostics Check		OK	09:39	

Data Download Date/Time: 12/30/2019 15:44

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARLOW,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: PATRICK,KENNETH
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 12/01/2019 02:31:19

Diagnostics Check		OK	02:22	
Air Blank		0.000	02:23	
Control Test		0.080	02:23	
Air Blank		0.000	02:24	
Subject Sample	# 1		02:27	No Sample Provided
Breath Volume		0.000	02:27	
Air Blank		0.000	02:27	
Air Blank		0.000	02:29	
Subject Sample	# 2		02:29	Subject Test Refused
Breath Volume		0.000	02:29	
Air Blank		0.000	02:30	
Control Test		0.081	02:30	
Air Blank		0.000	02:31	
Diagnostics Check		OK	02:31	

Data Download Date/Time: 12/15/2019 13:38

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARLOW,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: BELL,DAKOTA
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: WILKEN

DATE/TIME OF TEST: 12/01/2019 03:30:34

Diagnostics Check		OK	03:21	
Air Blank		0.000	03:22	
Control Test		0.080	03:22	
Air Blank		0.000	03:23	
Subject Sample	# 1		03:26	No Sample Provided
Breath Volume		0.000	03:26	
Air Blank		0.000	03:26	
Air Blank		0.000	03:28	
Subject Sample	# 2		03:29	Subject Test Refused
Breath Volume		0.000	03:29	
Air Blank		0.000	03:29	
Control Test		0.080	03:29	
Air Blank		0.000	03:30	
Diagnostics Check		OK	03:30	

Data Download Date/Time: 12/15/2019 13:38

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARLOW,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: ADAMS,STEVEN
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: BLEDSOE

DATE/TIME OF TEST: 12/05/2019 03:27:56

Diagnostics Check		OK	03:21
Air Blank		0.000	03:21
Control Test		0.081	03:22
Air Blank		0.000	03:22
Subject Sample	# 1	0.000	03:23
Breath Volume		2.218	03:23
Air Blank		0.000	03:24
Air Blank		0.000	03:26
Subject Sample	# 2	0.000	03:26
Breath Volume		2.367	03:26
Air Blank		0.000	03:27
Control Test		0.080	03:27
Air Blank		0.000	03:27
Diagnostics Check		OK	03:27

Data Download Date/Time: 12/15/2019 13:38

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARWOOD,CHARLES,M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: ADAMS,S
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:23

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 12/09/2019 22:53:42

Diagnostics Check		OK	22:47
Air Blank		0.000	22:47
Control Test		0.080	22:48
Air Blank		0.000	22:48
Subject Sample	# 1	0.000	22:49
Breath Volume		1.281	22:49
Air Blank		0.000	22:49
Air Blank		0.000	22:51
Subject Sample	# 2	0.000	22:52
Breath Volume		2.207	22:52
Air Blank		0.000	22:52
Control Test		0.079	22:53
Air Blank		0.000	22:53
Diagnostics Check		OK	22:53

Data Download Date/Time: 12/15/2019 13:38

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARLOW,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: FORTIN,BRETT
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: LOCKHART

DATE/TIME OF TEST: 12/15/2019 04:12:51

Diagnostics Check		OK	04:02	
Air Blank		0.000	04:02	
Control Test		0.080	04:03	
Air Blank		0.000	04:03	
Subject Sample	# 1		04:06	No Sample Provided
Breath Volume		0.000	04:06	
Air Blank		0.000	04:07	
Air Blank		0.000	04:09	
Subject Sample	# 2		04:11	Subject Test Refused
Breath Volume		0.000	04:11	
Air Blank		0.000	04:11	
Control Test		0.080	04:12	
Air Blank		0.000	04:12	
Diagnostics Check		OK	04:12	

Data Download Date/Time: 12/15/2019 13:38

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DANELLA,E,G
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JONES,T
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: ADMIN
ARREST TIME: 15:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:30

SUBJECT'S LAST NAME: LAYFIELD

DATE/TIME OF TEST: 12/16/2019 16:02:24

Diagnostics Check		OK	15:56
Air Blank		0.000	15:56
Control Test		0.080	15:57
Air Blank		0.000	15:57
Subject Sample	# 1	0.000	15:58
Breath Volume		1.835	15:58
Air Blank		0.000	15:58
Air Blank		0.000	16:00
Subject Sample	# 2	0.000	16:01
Breath Volume		1.707	16:01
Air Blank		0.000	16:01
Control Test		0.079	16:01
Air Blank		0.000	16:02
Diagnostics Check		OK	16:02

Data Download Date/Time: 01/27/2020 08:47

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARLOW,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: ADAMS,STEPHEN
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: KITE

DATE/TIME OF TEST: 12/19/2019 02:13:48

Diagnostics Check		OK	02:05	
Air Blank		0.000	02:05	
Control Test		0.081	02:05	
Air Blank		0.000	02:06	
Subject Sample	# 1		02:09	No Sample Provided
Breath Volume		0.000	02:09	
Air Blank		0.000	02:09	
Air Blank		0.000	02:11	
Subject Sample	# 2		02:12	Subject Test Refused
Breath Volume		0.000	02:12	
Air Blank		0.000	02:12	
Control Test		0.082	02:13	
Air Blank		0.000	02:13	
Diagnostics Check		OK	02:13	

Data Download Date/Time: 01/27/2020 08:47

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARLOW,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: BAUMGARTNER,NATHANIAL
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: STONE

DATE/TIME OF TEST: 12/19/2019 22:29:53

Diagnostics Check		OK	22:21	
Air Blank		0.000	22:21	
Control Test		0.080	22:22	
Air Blank		0.000	22:22	
Subject Sample	# 1		22:25	No Sample Provided
Breath Volume		0.000	22:25	
Air Blank		0.000	22:26	
Air Blank		0.000	22:28	
Subject Sample	# 2		22:28	Subject Test Refused
Breath Volume		0.000	22:28	
Air Blank		0.000	22:28	
Control Test		0.081	22:29	
Air Blank		0.000	22:29	
Diagnostics Check		OK	22:29	

Data Download Date/Time: 01/27/2020 08:47

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTSTATT,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: CAPORAL

DATE/TIME OF TEST: 12/26/2019 01:22:22

Diagnostics Check		OK	01:11	
Air Blank		0.000	01:11	
Control Test		0.080	01:11	
Air Blank		0.000	01:12	
Subject Sample	# 1		01:15	No Sample Provided
Breath Volume		0.000	01:15	
Air Blank		0.000	01:15	
Air Blank		0.000	01:17	
Subject Sample	# 2		01:21	No Sample Provided
Breath Volume		0.000	01:21	
Air Blank		0.000	01:21	
Control Test		0.080	01:21	
Air Blank		0.000	01:22	
Diagnostics Check		OK	01:22	

Data Download Date/Time: 01/27/2020 08:47

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTSTATT,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:49

SUBJECT'S LAST NAME: SNYDER

DATE/TIME OF TEST: 12/26/2019 05:19:12

Diagnostics Check		OK	05:12
Air Blank		0.000	05:12
Control Test		0.080	05:13
Air Blank		0.000	05:13
Subject Sample	# 1	0.115	05:14
Breath Volume		3.015	05:14
Air Blank		0.000	05:15
Air Blank		0.000	05:17
Subject Sample	# 2	0.112	05:17
Breath Volume		2.355	05:17
Air Blank		0.000	05:18
Control Test		0.080	05:18
Air Blank		0.000	05:19
Diagnostics Check		OK	05:19

Data Download Date/Time: 01/27/2020 08:47

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNOW,C,D
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: SMITH,J
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 10:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:05

SUBJECT'S LAST NAME: TOBLER

DATE/TIME OF TEST: 12/29/2019 11:35:24

Diagnostics Check		OK	11:29
Air Blank		0.000	11:29
Control Test		0.079	11:29
Air Blank		0.000	11:30
Subject Sample	# 1	0.226	11:30
Breath Volume		2.023	11:30
Air Blank		0.000	11:31
Air Blank		0.000	11:33
Subject Sample	# 2	0.229	11:33
Breath Volume		2.230	11:33
Air Blank		0.000	11:34
Control Test		0.079	11:34
Air Blank		0.000	11:35
Diagnostics Check		OK	11:35

Data Download Date/Time: 01/27/2020 08:47

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEASE,MATTHEW,W
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: SHAFFER,ANTHONY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 11/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181622
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 12/24/2019 19:09:04

Diagnostics Check		OK	19:01
Air Blank		0.000	19:01
Control Test		0.082	19:02
Air Blank		0.000	19:02
Subject Sample	# 1	0.000	19:04
Breath Volume		1.796	19:04
Air Blank		0.000	19:04
Air Blank		0.000	19:06
Subject Sample	# 2	0.000	19:07
Breath Volume		1.933	19:07
Air Blank		0.000	19:08
Control Test		0.082	19:08
Air Blank		0.000	19:08
Diagnostics Check		OK	19:09

Data Download Date/Time: 12/31/2019 15:24

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENNELL,RUSSELL,M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: IVEY,BRIAN
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 11/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: AIKEN

DATE/TIME OF TEST: 12/26/2019 22:04:43

Diagnostics Check		OK	21:56
Air Blank		0.000	21:56
Control Test		0.081	21:57
Air Blank		0.000	21:57
Subject Sample	# 1	0.151	21:58
Breath Volume		1.125	21:58
Air Blank		0.000	21:59
Air Blank		0.000	22:00
Subject Sample	# 2	0.167	22:03
Breath Volume		1.859	22:03
Air Blank		0.000	22:03
Control Test		0.080	22:04
Air Blank		0.000	22:04
Diagnostics Check		OK	22:04

Data Download Date/Time: 12/31/2019 15:24

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ;6360100456578077164,SUNIL,
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: KENNY,J
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 11/25/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 12:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: DEONARINE

DATE/TIME OF TEST: 12/11/2019 01:34:52

Diagnostics Check		OK	01:30	
Air Blank		0.000	01:30	
Control Test		0.079	01:31	
Air Blank		0.000	01:31	
Subject Sample	# 1	0.135	01:33	
Breath Volume		1.187	01:33	
Air Blank			01:34	Purge Fail
Air Blank			01:34	Purge Fail

Data Download Date/Time: 12/19/2019 11:13

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDES,D,
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: KENNY,J
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 11/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 12:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: DEONARINE

DATE/TIME OF TEST: 12/11/2019 01:45:46

Diagnostics Check		OK	01:36	
Air Blank		0.000	01:36	
Control Test		0.079	01:37	
Air Blank		0.000	01:37	
Subject Sample	# 1	0.111	01:40	Volume Not Met
Breath Volume		0.519	01:40	
Air Blank		0.000	01:41	
Air Blank		0.000	01:43	
Subject Sample	# 2		01:44	Subject Test Refused
Breath Volume		0.433	01:44	
Air Blank		0.000	01:44	
Control Test		0.077	01:45	
Air Blank		0.000	01:45	
Diagnostics Check		OK	01:45	

Data Download Date/Time: 12/19/2019 11:13

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WETHERINGTON,STEPHEN,L
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: WETHERINGTON,STEPHEN
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 11/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: DESINAT

DATE/TIME OF TEST: 12/15/2019 04:01:42

Diagnostics Check		OK	03:55
Air Blank		0.000	03:55
Control Test		0.079	03:55
Air Blank		0.000	03:56
Subject Sample	# 1	0.130	03:57
Breath Volume		2.535	03:57
Air Blank		0.000	03:57
Air Blank		0.000	03:59
Subject Sample	# 2	0.131	04:00
Breath Volume		2.679	04:00
Air Blank		0.000	04:00
Control Test		0.079	04:01
Air Blank		0.000	04:01
Diagnostics Check		OK	04:01

Data Download Date/Time: 12/19/2019 11:13

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WETHERINGTON,STEPHEN,L
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: ATRILES,LUIS
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 11/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: LEFLER

DATE/TIME OF TEST: 12/15/2019 04:32:09

Diagnostics Check		OK	04:24
Air Blank		0.000	04:24
Control Test		0.080	04:25
Air Blank		0.000	04:25
Subject Sample	# 1	0.110	04:26
Breath Volume		2.429	04:26
Air Blank		0.000	04:27
Air Blank		0.000	04:29
Subject Sample	# 2	0.108	04:30
Breath Volume		2.640	04:30
Air Blank		0.000	04:31
Control Test		0.082	04:31
Air Blank		0.000	04:32
Diagnostics Check		OK	04:32

Data Download Date/Time: 12/19/2019 11:13

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENE,KYLE,T
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: BULMAN,C
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 11/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: MCCARTHY

DATE/TIME OF TEST: 12/15/2019 05:06:41

Diagnostics Check		OK	04:49	
Air Blank		0.000	04:50	
Control Test		0.079	04:50	
Air Blank		0.000	04:51	
Subject Sample	# 1	0.165	04:54	
Breath Volume		1.253	04:54	
Air Blank		0.000	04:55	
Air Blank		0.000	04:56	
Subject Sample	# 2	0.145	04:59	Volume Not Met
Breath Volume		0.019	04:59	
Air Blank		0.000	05:00	
Air Blank		0.000	05:02	
Subject Sample	# 3	0.158	05:05	
Breath Volume		1.308	05:05	
Air Blank		0.000	05:05	
Control Test		0.077	05:06	
Air Blank		0.000	05:06	
Diagnostics Check		OK	05:06	

Data Download Date/Time: 12/19/2019 11:13

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WETHERINGTON,STEPHEN,L
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: SCHWARTZ,TYSON
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 11/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:47

SUBJECT'S LAST NAME: POULIN

DATE/TIME OF TEST: 12/15/2019 21:26:15

Diagnostics Check		OK	21:14
Air Blank		0.000	21:14
Control Test		0.079	21:15
Air Blank		0.000	21:15
Subject Sample	# 1	0.206	21:16
Breath Volume		1.191	21:16
Air Blank		0.000	21:17
Air Blank		0.000	21:19
Subject Sample	# 2	0.240	21:21
Breath Volume		2.691	21:21
Air Blank		0.000	21:21
Air Blank		0.000	21:23
Subject Sample	# 3	0.227	21:24
Breath Volume		2.273	21:24
Air Blank		0.000	21:25
Control Test		0.078	21:25
Air Blank		0.000	21:26
Diagnostics Check		OK	21:26

Data Download Date/Time: 12/19/2019 11:13

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARA-VALDERRAMA,MARIO,A
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: VARA-VALDERRAMA,MARIO
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 12/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 14:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:52

SUBJECT'S LAST NAME: GUZMAN

DATE/TIME OF TEST: 12/23/2019 15:41:57

Diagnostics Check		OK	15:35
Air Blank		0.000	15:36
Control Test		0.079	15:36
Air Blank		0.000	15:36
Subject Sample	# 1	0.349	15:37
Breath Volume		2.984	15:37
Air Blank		0.000	15:38
Air Blank		0.000	15:39
Subject Sample	# 2	0.336	15:40
Breath Volume		2.507	15:40
Air Blank		0.000	15:41
Control Test		0.078	15:41
Air Blank		0.000	15:41
Diagnostics Check		OK	15:41

Data Download Date/Time: 01/29/2020 14:09

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCULLOUGH,SEAN,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LESPERANCE,DANIEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 12/01/2019 04:20:07

Diagnostics Check		OK	04:13
Air Blank		0.000	04:13
Control Test		0.080	04:14
Air Blank		0.000	04:14
Subject Sample	# 1	0.136	04:15
Breath Volume		1.843	04:15
Air Blank		0.000	04:15
Air Blank		0.000	04:17
Subject Sample	# 2	0.131	04:18
Breath Volume		1.761	04:18
Air Blank		0.000	04:19
Control Test		0.080	04:19
Air Blank		0.000	04:20
Diagnostics Check		OK	04:20

Data Download Date/Time: 12/23/2019 10:29

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOWRY,JAMIE,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SMITH,EMILY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: PIRRONE

DATE/TIME OF TEST: 12/01/2019 20:22:50

Diagnostics Check		OK	20:15	
Air Blank		0.000	20:16	
Control Test		0.080	20:16	
Air Blank		0.000	20:17	
Subject Sample	# 1	0.221	20:17	
Breath Volume		1.820	20:17	
Air Blank		0.000	20:18	
Air Blank		0.000	20:20	
Subject Sample	# 2		20:21	Subject Test Refused
Breath Volume		2.449	20:21	
Air Blank		0.000	20:21	
Control Test		0.079	20:22	
Air Blank		0.000	20:22	
Diagnostics Check		OK	20:22	

Data Download Date/Time: 12/23/2019 10:29

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARRETT,DAVID,S
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,KRISTEN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: RESTO

DATE/TIME OF TEST: 12/04/2019 00:32:26

Diagnostics Check		OK	00:28	
Air Blank		0.000	00:28	
Control Test		0.080	00:29	
Air Blank		0.000	00:29	
Subject Sample	# 1		00:31	Subject Test Refused
Breath Volume		0.000	00:31	
Air Blank		0.000	00:31	
Control Test		0.081	00:31	
Air Blank		0.000	00:32	
Diagnostics Check		OK	00:32	

Data Download Date/Time: 12/23/2019 10:29

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LESPERANCE,DANIEL,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LAMOUR,ZACHARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: VASSELL

DATE/TIME OF TEST: 12/04/2019 21:36:07

Diagnostics Check		OK	21:29
Air Blank		0.000	21:30
Control Test		0.080	21:30
Air Blank		0.000	21:30
Subject Sample	# 1	0.077	21:31
Breath Volume		1.992	21:31
Air Blank		0.000	21:31
Air Blank		0.000	21:33
Subject Sample	# 2	0.075	21:34
Breath Volume		2.097	21:34
Air Blank		0.000	21:35
Control Test		0.081	21:35
Air Blank		0.000	21:36
Diagnostics Check		OK	21:36

Data Download Date/Time: 12/23/2019 10:29

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LESPERANCE,DANIEL,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCCULLOUGH,SEAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 12/05/2019 01:02:48

Diagnostics Check		OK	00:51	
Air Blank		0.000	00:51	
Control Test		0.080	00:52	
Air Blank		0.000	00:52	
Subject Sample	# 1		00:55	No Sample Provided
Breath Volume		0.000	00:55	
Air Blank		0.000	00:56	
Air Blank		0.000	00:58	
Subject Sample	# 2		01:01	No Sample Provided
Breath Volume		0.000	01:01	
Air Blank		0.000	01:01	
Control Test		0.080	01:02	
Air Blank		0.000	01:02	
Diagnostics Check		OK	01:02	

Data Download Date/Time: 12/23/2019 10:29

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAMOUR,ZACHARY,T
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HINES,PAUL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 12/05/2019 01:18:48

Diagnostics Check		OK	01:10	
Air Blank		0.000	01:10	
Control Test		0.080	01:10	
Air Blank		0.000	01:11	
Subject Sample	# 1		01:14	No Sample Provided
Breath Volume		0.000	01:14	
Air Blank		0.000	01:15	
Air Blank		0.000	01:17	
Subject Sample	# 2		01:17	RFI Detect
Breath Volume		0.000	01:17	
Air Blank		0.000	01:17	
Control Test		0.081	01:18	
Air Blank		0.000	01:18	
Diagnostics Check		OK	01:18	

Data Download Date/Time: 12/23/2019 10:29

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAMOUR,ZACHARY,T
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HINES,PAUL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 12/05/2019 01:26:20

Diagnostics Check		OK	01:20	
Air Blank		0.000	01:20	
Control Test		0.080	01:20	
Air Blank		0.000	01:21	
Subject Sample	# 1	0.144	01:22	
Breath Volume		2.289	01:22	
Air Blank		0.000	01:23	
Air Blank		0.000	01:25	
Subject Sample	# 2	0.139	01:25	
Breath Volume		2.109	01:25	
Air Blank			01:25	RFI Detect
Air Blank		0.000	01:26	

Data Download Date/Time: 12/23/2019 10:29

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAMOUR,ZACHARY,T
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HINES,PAUL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 12/05/2019 01:37:01

Diagnostics Check		OK	01:30
Air Blank		0.000	01:31
Control Test		0.081	01:31
Air Blank		0.000	01:31
Subject Sample	# 1	0.139	01:32
Breath Volume		1.667	01:32
Air Blank		0.000	01:33
Air Blank		0.000	01:35
Subject Sample	# 2	0.145	01:35
Breath Volume		1.480	01:35
Air Blank		0.000	01:36
Control Test		0.079	01:36
Air Blank		0.000	01:36
Diagnostics Check		OK	01:37

Data Download Date/Time: 12/23/2019 10:29

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARRETT,DAVID,S
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: POST,ALEXANDER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: LANEY

DATE/TIME OF TEST: 12/05/2019 04:38:46

Diagnostics Check		OK	04:32
Air Blank		0.000	04:32
Control Test		0.080	04:32
Air Blank		0.000	04:33
Subject Sample	# 1	0.000	04:34
Breath Volume		2.808	04:34
Air Blank		0.000	04:34
Air Blank		0.000	04:36
Subject Sample	# 2	0.000	04:37
Breath Volume		3.089	04:37
Air Blank		0.000	04:37
Control Test		0.080	04:38
Air Blank		0.000	04:38
Diagnostics Check		OK	04:38

Data Download Date/Time: 12/23/2019 10:29

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LESPERANCE,DANIEL,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CRUM,ZACHARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: DANIELS

DATE/TIME OF TEST: 12/05/2019 21:58:14

Diagnostics Check		OK	21:50
Air Blank		0.000	21:51
Control Test		0.080	21:51
Air Blank		0.000	21:52
Subject Sample	# 1	0.134	21:53
Breath Volume		2.332	21:53
Air Blank		0.000	21:53
Air Blank		0.000	21:55
Subject Sample	# 2	0.129	21:56
Breath Volume		2.058	21:56
Air Blank		0.000	21:57
Control Test		0.079	21:57
Air Blank		0.000	21:58
Diagnostics Check		OK	21:58

Data Download Date/Time: 12/23/2019 10:29

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAMOUR,ZACHARY,T
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCCULLOUGH,SEAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: THRUMSTON

DATE/TIME OF TEST: 12/06/2019 01:31:20

Diagnostics Check		OK	01:23
Air Blank		0.000	01:24
Control Test		0.080	01:24
Air Blank		0.000	01:25
Subject Sample	# 1	0.117	01:26
Breath Volume		1.570	01:26
Air Blank		0.000	01:27
Air Blank		0.000	01:29
Subject Sample	# 2	0.115	01:29
Breath Volume		1.640	01:29
Air Blank		0.000	01:30
Control Test		0.080	01:30
Air Blank		0.000	01:31
Diagnostics Check		OK	01:31

Data Download Date/Time: 12/23/2019 10:30

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LESPERANCE,DANIEL,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,KRISTEN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:37

SUBJECT'S LAST NAME: TERGESEN

DATE/TIME OF TEST: 12/06/2019 20:03:47

Diagnostics Check		OK	19:57
Air Blank		0.000	19:58
Control Test		0.081	19:58
Air Blank		0.000	19:59
Subject Sample	# 1	0.000	19:59
Breath Volume		2.453	19:59
Air Blank		0.000	20:00
Air Blank		0.000	20:02
Subject Sample	# 2	0.000	20:02
Breath Volume		2.214	20:02
Air Blank		0.000	20:02
Control Test		0.080	20:03
Air Blank		0.000	20:03
Diagnostics Check		OK	20:03

Data Download Date/Time: 12/23/2019 10:30

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINES,PAUL,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LAMOUR,ZACHARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: ALEXANDER

DATE/TIME OF TEST: 12/07/2019 00:42:29

Diagnostics Check		OK	00:36
Air Blank		0.000	00:36
Control Test		0.081	00:37
Air Blank		0.000	00:37
Subject Sample	# 1	0.093	00:38
Breath Volume		2.230	00:38
Air Blank		0.000	00:38
Air Blank		0.000	00:40
Subject Sample	# 2	0.092	00:41
Breath Volume		2.238	00:41
Air Blank		0.000	00:41
Control Test		0.080	00:41
Air Blank		0.000	00:42
Diagnostics Check		OK	00:42

Data Download Date/Time: 12/23/2019 10:30

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUM,ZACKARY,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCCULLOUGH,SEAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 12/07/2019 03:09:45

Diagnostics Check		OK	03:03
Air Blank		0.000	03:03
Control Test		0.079	03:04
Air Blank		0.000	03:04
Subject Sample	# 1	0.083	03:05
Breath Volume		1.542	03:05
Air Blank		0.000	03:05
Air Blank		0.000	03:07
Subject Sample	# 2	0.080	03:08
Breath Volume		1.742	03:08
Air Blank		0.000	03:08
Control Test		0.080	03:09
Air Blank		0.000	03:09
Diagnostics Check		OK	03:09

Data Download Date/Time: 12/23/2019 10:30

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LESPERANCE,DANIEL,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HINES,PAUL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: ETHERIDGE

DATE/TIME OF TEST: 12/07/2019 04:20:07

Diagnostics Check		OK	04:13
Air Blank		0.000	04:14
Control Test		0.080	04:14
Air Blank		0.000	04:15
Subject Sample	# 1	0.190	04:15
Breath Volume		1.914	04:15
Air Blank		0.000	04:16
Air Blank		0.000	04:18
Subject Sample	# 2	0.181	04:18
Breath Volume		1.468	04:18
Air Blank		0.000	04:19
Control Test		0.080	04:19
Air Blank		0.000	04:20
Diagnostics Check		OK	04:20

Data Download Date/Time: 12/23/2019 10:30

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAIA,MICHAEL,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SAIA,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: MARKER

DATE/TIME OF TEST: 12/08/2019 00:49:42

Diagnostics Check		OK	00:44	
Air Blank		0.000	00:44	
Control Test		0.080	00:45	
Air Blank		0.000	00:45	
Subject Sample	# 1		00:48	Subject Test Refused
Breath Volume		0.000	00:48	
Air Blank		0.000	00:48	
Control Test		0.079	00:49	
Air Blank		0.000	00:49	
Diagnostics Check		OK	00:49	

Data Download Date/Time: 12/23/2019 10:30

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAMOUR,ZACHARY,T
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LESPERANCE,DANIEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:19

SUBJECT'S LAST NAME: QUIGLEY

DATE/TIME OF TEST: 12/11/2019 20:47:30

Diagnostics Check		OK	20:40
Air Blank		0.000	20:40
Control Test		0.081	20:41
Air Blank		0.000	20:41
Subject Sample	# 1	0.140	20:42
Breath Volume		3.351	20:42
Air Blank		0.000	20:43
Air Blank		0.000	20:45
Subject Sample	# 2	0.140	20:46
Breath Volume		3.039	20:46
Air Blank		0.000	20:46
Control Test		0.080	20:46
Air Blank		0.000	20:47
Diagnostics Check		OK	20:47

Data Download Date/Time: 12/23/2019 10:30

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINES,PAUL,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CRUM,ZACKARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: BECKMANN

DATE/TIME OF TEST: 12/11/2019 22:24:21

Diagnostics Check		OK	22:12	
Air Blank		0.000	22:13	
Control Test		0.080	22:13	
Air Blank		0.000	22:14	
Subject Sample	# 1		22:17	No Sample Provided
Breath Volume		0.000	22:17	
Air Blank		0.000	22:17	
Air Blank		0.000	22:19	
Subject Sample	# 2		22:22	No Sample Provided
Breath Volume		0.000	22:22	
Air Blank		0.000	22:23	
Control Test		0.081	22:23	
Air Blank		0.000	22:24	
Diagnostics Check		OK	22:24	

Data Download Date/Time: 12/23/2019 10:30

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINES,PAUL,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LAMOUR,ZACHARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:14

SUBJECT'S LAST NAME: SAUVAGEOT

DATE/TIME OF TEST: 12/12/2019 17:40:49

Diagnostics Check		OK	17:34
Air Blank		0.000	17:35
Control Test		0.080	17:35
Air Blank		0.000	17:35
Subject Sample	# 1	0.000	17:36
Breath Volume		3.097	17:36
Air Blank		0.000	17:36
Air Blank		0.000	17:38
Subject Sample	# 2	0.000	17:39
Breath Volume		2.753	17:39
Air Blank		0.000	17:39
Control Test		0.080	17:40
Air Blank		0.000	17:40
Diagnostics Check		OK	17:40

Data Download Date/Time: 12/23/2019 10:30

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LESPERANCE,DANIEL,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCKENZIE,RONALD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 11:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:18

SUBJECT'S LAST NAME: ROMERO

DATE/TIME OF TEST: 12/13/2019 11:51:36

Diagnostics Check		OK	11:45
Air Blank		0.000	11:45
Control Test		0.079	11:46
Air Blank		0.000	11:46
Subject Sample	# 1	0.261	11:47
Breath Volume		1.808	11:47
Air Blank		0.000	11:47
Air Blank		0.000	11:49
Subject Sample	# 2	0.271	11:50
Breath Volume		1.531	11:50
Air Blank		0.000	11:50
Control Test		0.080	11:51
Air Blank		0.000	11:51
Diagnostics Check		OK	11:51

Data Download Date/Time: 12/23/2019 10:30

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUM,ZACKARY,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCCULLOUGH,SEAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: PETRILL

DATE/TIME OF TEST: 12/13/2019 23:50:53

Diagnostics Check		OK	23:44
Air Blank		0.000	23:45
Control Test		0.079	23:45
Air Blank		0.000	23:46
Subject Sample	# 1	0.098	23:46
Breath Volume		1.820	23:46
Air Blank		0.000	23:47
Air Blank		0.000	23:48
Subject Sample	# 2	0.099	23:49
Breath Volume		1.203	23:49
Air Blank		0.000	23:49
Control Test		0.079	23:50
Air Blank		0.000	23:50
Diagnostics Check		OK	23:50

Data Download Date/Time: 12/23/2019 10:30

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCULLOUGH,SEAN,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HINES,PAUL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: RANSOM

DATE/TIME OF TEST: 12/14/2019 00:47:24

Diagnostics Check		OK	00:41
Air Blank		0.000	00:41
Control Test		0.080	00:41
Air Blank		0.000	00:42
Subject Sample	# 1	0.213	00:42
Breath Volume		1.750	00:42
Air Blank		0.000	00:43
Air Blank		0.000	00:45
Subject Sample	# 2	0.209	00:45
Breath Volume		2.980	00:45
Air Blank		0.000	00:46
Control Test		0.080	00:46
Air Blank		0.000	00:47
Diagnostics Check		OK	00:47

Data Download Date/Time: 12/23/2019 10:30

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAMOUR,ZACHARY,T
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ADAMS,DWAYNE
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: GRUSCHKE

DATE/TIME OF TEST: 12/14/2019 22:12:24

Diagnostics Check		OK	22:05
Air Blank		0.000	22:05
Control Test		0.080	22:05
Air Blank		0.000	22:06
Subject Sample	# 1	0.279	22:07
Breath Volume		5.464	22:07
Air Blank		0.000	22:08
Air Blank		0.000	22:10
Subject Sample	# 2	0.269	22:10
Breath Volume		5.562	22:10
Air Blank		0.000	22:11
Control Test		0.078	22:11
Air Blank		0.000	22:12
Diagnostics Check		OK	22:12

Data Download Date/Time: 12/23/2019 10:30

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAMOUR,ZACHARY,T
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CRUM,ZACKARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: MANDERSCHEID

DATE/TIME OF TEST: 12/14/2019 23:08:41

Diagnostics Check		OK	23:05	
Air Blank		0.000	23:06	
Control Test		0.080	23:06	
Air Blank		0.000	23:07	
Subject Sample	# 1		23:07	Subject Test Refused
Breath Volume		0.000	23:07	
Air Blank		0.000	23:07	
Control Test		0.081	23:08	
Air Blank		0.000	23:08	
Diagnostics Check		OK	23:08	

Data Download Date/Time: 12/23/2019 10:30

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCULLOUGH,SEAN,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HINES,PAUL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: ANDERSEN

DATE/TIME OF TEST: 12/15/2019 04:35:40

Diagnostics Check		OK	04:28
Air Blank		0.000	04:29
Control Test		0.081	04:29
Air Blank		0.000	04:30
Subject Sample	# 1	0.093	04:31
Breath Volume		2.019	04:31
Air Blank		0.000	04:31
Air Blank		0.000	04:33
Subject Sample	# 2	0.088	04:34
Breath Volume		2.199	04:34
Air Blank		0.000	04:34
Control Test		0.080	04:35
Air Blank		0.000	04:35
Diagnostics Check		OK	04:35

Data Download Date/Time: 12/23/2019 10:31

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABELN,SHON,W
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: ABELN,SHON
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: COURT ORD
ARREST TIME: 11:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:30

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 12/02/2019 12:00:58

Diagnostics Check		OK	11:55
Air Blank		0.000	11:55
Control Test		0.080	11:55
Air Blank		0.000	11:56
Subject Sample	# 1	0.000	11:56
Breath Volume		1.953	11:56
Air Blank		0.000	11:57
Air Blank		0.000	11:59
Subject Sample	# 2	0.000	11:59
Breath Volume		1.796	11:59
Air Blank		0.000	12:00
Control Test		0.080	12:00
Air Blank		0.000	12:00
Diagnostics Check		OK	12:00

Data Download Date/Time: 12/30/2019 20:23

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINISH,BARRY,A
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: BOGUSKY,DENNIS
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: LAMB

DATE/TIME OF TEST: 12/05/2019 04:59:19

Diagnostics Check		OK	04:52
Air Blank		0.000	04:53
Control Test		0.081	04:53
Air Blank		0.000	04:53
Subject Sample	# 1	0.083	04:54
Breath Volume		1.371	04:54
Air Blank		0.000	04:55
Air Blank		0.000	04:57
Subject Sample	# 2	0.086	04:57
Breath Volume		1.265	04:57
Air Blank		0.000	04:58
Control Test		0.080	04:58
Air Blank		0.000	04:59
Diagnostics Check		OK	04:59

Data Download Date/Time: 12/30/2019 20:23

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBLES,J,V
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: BURRESS,R
ARRESTING OFFICER AGENCY: Lake Placid PD
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:03

SUBJECT'S LAST NAME: BUIS II

DATE/TIME OF TEST: 12/08/2019 16:26:12

Diagnostics Check		OK	16:23	
Air Blank		0.000	16:23	
Control Test		0.080	16:23	
Air Blank		0.000	16:24	
Subject Sample	# 1		16:24	Subject Test Refused
Breath Volume		0.000	16:24	
Air Blank		0.000	16:25	
Control Test		0.080	16:25	
Air Blank		0.000	16:26	
Diagnostics Check		OK	16:26	

Data Download Date/Time: 12/30/2019 20:23

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINISH,BARRY,A
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SALVADOR,ELDER
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: CROWE

DATE/TIME OF TEST: 12/14/2019 23:11:17

Diagnostics Check		OK	23:07	
Air Blank		0.000	23:07	
Control Test		0.080	23:07	
Air Blank		0.000	23:08	
Subject Sample	# 1		23:09	Subject Test Refused
Breath Volume		0.000	23:09	
Air Blank		0.000	23:10	
Control Test		0.080	23:10	
Air Blank		0.000	23:11	
Diagnostics Check		OK	23:11	

Data Download Date/Time: 12/30/2019 20:23

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEEFE,JAYME,W
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: GARRISON,JOHN
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 12/15/2019 00:11:28

Diagnostics Check		OK	00:04
Air Blank		0.000	00:04
Control Test		0.080	00:05
Air Blank		0.000	00:05
Subject Sample	# 1	0.213	00:06
Breath Volume		1.496	00:06
Air Blank		0.000	00:07
Air Blank		0.000	00:09
Subject Sample	# 2	0.206	00:09
Breath Volume		1.285	00:09
Air Blank		0.000	00:10
Control Test		0.080	00:10
Air Blank		0.000	00:11
Diagnostics Check		OK	00:11

Data Download Date/Time: 12/30/2019 20:23

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GLASCO,GREGORY,S
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: BOGUSKY,DENNIS
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: JARVIS

DATE/TIME OF TEST: 12/15/2019 01:16:45

Diagnostics Check		OK	01:10
Air Blank		0.000	01:10
Control Test		0.079	01:11
Air Blank		0.000	01:11
Subject Sample	# 1	0.189	01:12
Breath Volume		1.125	01:12
Air Blank		0.000	01:13
Air Blank		0.000	01:14
Subject Sample	# 2	0.183	01:15
Breath Volume		1.382	01:15
Air Blank		0.000	01:15
Control Test		0.079	01:16
Air Blank		0.000	01:16
Diagnostics Check		OK	01:16

Data Download Date/Time: 12/30/2019 20:23

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARLOW,ROYCE,B
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: BOGUSKY,DENNIS
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:25

SUBJECT'S LAST NAME: ROMERO

DATE/TIME OF TEST: 12/15/2019 07:53:31

Diagnostics Check		OK	07:46
Air Blank		0.000	07:47
Control Test		0.080	07:47
Air Blank		0.000	07:48
Subject Sample	# 1	0.029	07:48
Breath Volume		1.593	07:48
Air Blank		0.000	07:49
Air Blank		0.000	07:51
Subject Sample	# 2	0.030	07:52
Breath Volume		1.855	07:52
Air Blank		0.000	07:52
Control Test		0.082	07:52
Air Blank		0.000	07:53
Diagnostics Check		OK	07:53

Data Download Date/Time: 12/30/2019 20:23

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGOODIAL,TRISTON,T
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: MCLAUGHLIN,TERENCE
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: NEESE

DATE/TIME OF TEST: 12/17/2019 05:17:30

Diagnostics Check		OK	05:14	
Air Blank		0.000	05:15	
Control Test		0.078	05:15	
Air Blank		0.000	05:15	
Subject Sample	# 1		05:16	Subject Test Refused
Breath Volume		0.000	05:16	
Air Blank		0.000	05:16	
Control Test		0.078	05:16	
Air Blank		0.000	05:17	
Diagnostics Check		OK	05:17	

Data Download Date/Time: 12/30/2019 20:23

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABELN,C,S
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SALVADOR,E
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 12/23/2019 02:39:28

Diagnostics Check		OK	02:32
Air Blank		0.000	02:33
Control Test		0.079	02:33
Air Blank		0.000	02:34
Subject Sample	# 1	0.195	02:34
Breath Volume		2.554	02:34
Air Blank		0.000	02:35
Air Blank		0.000	02:37
Subject Sample	# 2	0.192	02:37
Breath Volume		1.910	02:37
Air Blank		0.000	02:38
Control Test		0.079	02:38
Air Blank		0.000	02:39
Diagnostics Check		OK	02:39

Data Download Date/Time: 12/30/2019 20:23

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,DUSTIN,J
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: MILLER,T
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 14:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:40

SUBJECT'S LAST NAME: KWITOWSKI

DATE/TIME OF TEST: 12/26/2019 15:07:12

Diagnostics Check		OK	15:00
Air Blank		0.000	15:01
Control Test		0.080	15:01
Air Blank		0.000	15:02
Subject Sample	# 1	0.000	15:02
Breath Volume		2.269	15:02
Air Blank		0.000	15:03
Air Blank		0.000	15:05
Subject Sample	# 2	0.000	15:05
Breath Volume		2.082	15:05
Air Blank		0.000	15:06
Control Test		0.080	15:06
Air Blank		0.000	15:07
Diagnostics Check		OK	15:07

Data Download Date/Time: 12/30/2019 20:23

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBLES,J,V
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: MCLAUGHLIN,T
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: GRAHAM

DATE/TIME OF TEST: 12/27/2019 03:19:06

Diagnostics Check		OK	03:15	
Air Blank		0.000	03:16	
Control Test		0.080	03:16	
Air Blank		0.000	03:17	
Subject Sample	# 1		03:17	Subject Test Refused
Breath Volume		0.000	03:17	
Air Blank		0.000	03:18	
Control Test		0.079	03:18	
Air Blank		0.000	03:19	
Diagnostics Check		OK	03:19	

Data Download Date/Time: 12/30/2019 20:23

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBLES,J,V
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: MARIN,A
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 12/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:31

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 12/31/2019 19:58:14

Diagnostics Check		OK	19:51
Air Blank		0.000	19:52
Control Test		0.081	19:52
Air Blank		0.000	19:53
Subject Sample	# 1	0.000	19:53
Breath Volume		3.367	19:53
Air Blank		0.000	19:54
Air Blank		0.000	19:56
Subject Sample	# 2	0.000	19:56
Breath Volume		3.050	19:56
Air Blank		0.000	19:57
Control Test		0.080	19:57
Air Blank		0.000	19:58
Diagnostics Check		OK	19:58

Data Download Date/Time: 01/13/2020 06:36

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELCHER,ALEX,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 12/01/2019 01:28:53

Diagnostics Check		OK	01:17	
Air Blank		0.000	01:17	
Control Test		0.080	01:18	
Air Blank		0.000	01:18	
Subject Sample	# 1		01:21	No Sample Provided
Breath Volume		0.000	01:21	
Air Blank		0.000	01:22	
Air Blank		0.000	01:24	
Subject Sample	# 2		01:27	No Sample Provided
Breath Volume		0.000	01:27	
Air Blank		0.000	01:27	
Control Test		0.080	01:28	
Air Blank		0.000	01:28	
Diagnostics Check		OK	01:28	

Data Download Date/Time: 12/26/2019 14:58

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COKE,CLEVELAND,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: PROPST,JOHN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: KEEFE

DATE/TIME OF TEST: 12/01/2019 03:40:06

Diagnostics Check		OK	03:33
Air Blank		0.000	03:34
Control Test		0.080	03:34
Air Blank		0.000	03:35
Subject Sample	# 1	0.146	03:35
Breath Volume		3.246	03:35
Air Blank		0.000	03:36
Air Blank		0.000	03:38
Subject Sample	# 2	0.137	03:38
Breath Volume		3.148	03:38
Air Blank		0.000	03:39
Control Test		0.079	03:39
Air Blank		0.000	03:40
Diagnostics Check		OK	03:40

Data Download Date/Time: 12/26/2019 14:58

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELCHER,ALEX,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: AVERA

DATE/TIME OF TEST: 12/01/2019 04:14:16

Diagnostics Check		OK	04:01	
Air Blank		0.000	04:02	
Control Test		0.080	04:02	
Air Blank		0.000	04:02	
Subject Sample	# 1	0.272	04:06	Volume Not Met
Breath Volume		1.039	04:06	
Air Blank		0.000	04:06	
Air Blank		0.000	04:08	
Subject Sample	# 2	0.279	04:09	
Breath Volume		1.121	04:09	
Air Blank		0.000	04:10	
Air Blank		0.000	04:11	
Subject Sample	# 3	0.269	04:12	
Breath Volume		1.347	04:12	
Air Blank		0.000	04:13	
Control Test		0.079	04:13	
Air Blank		0.000	04:14	
Diagnostics Check		OK	04:14	

Data Download Date/Time: 12/26/2019 14:58

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKSON,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: JACKSON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 11/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 14:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:50

SUBJECT'S LAST NAME: SUTTON CRAPPS

DATE/TIME OF TEST: 12/05/2019 16:14:02

Diagnostics Check		OK	16:06
Air Blank		0.000	16:06
Control Test		0.080	16:07
Air Blank		0.000	16:07
Subject Sample	# 1	0.148	16:08
Breath Volume		1.617	16:08
Air Blank		0.000	16:08
Air Blank		0.000	16:10
Subject Sample	# 2	0.148	16:12
Breath Volume		1.632	16:12
Air Blank		0.000	16:13
Control Test		0.079	16:13
Air Blank		0.000	16:13
Diagnostics Check		OK	16:14

Data Download Date/Time: 12/26/2019 14:58

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COKE,CLEVELAND,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SEYMOUR,KEVIN
ARRESTING OFFICER AGENCY: Auburndale PD
LAST AGENCY INSPECTION: 11/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: ELKINS

DATE/TIME OF TEST: 12/05/2019 23:17:47

Diagnostics Check		OK	23:07	
Air Blank		0.000	23:08	
Control Test		0.080	23:08	
Air Blank		0.000	23:09	
Subject Sample	# 1	0.163	23:12	Volume Not Met
Breath Volume		0.242	23:12	
Air Blank		0.000	23:13	
Air Blank		0.000	23:14	
Subject Sample	# 2		23:16	Subject Test Refused
Breath Volume		0.167	23:16	
Air Blank		0.000	23:16	
Control Test		0.078	23:17	
Air Blank		0.000	23:17	
Diagnostics Check		OK	23:17	

Data Download Date/Time: 12/26/2019 14:58

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUGAN,LAKEBRA,S
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SEYMOUR,KEVIN
ARRESTING OFFICER AGENCY: Auburndale PD
LAST AGENCY INSPECTION: 11/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: ELKINS

DATE/TIME OF TEST: 12/05/2019 23:48:59

Diagnostics Check		OK	23:39
Air Blank		0.000	23:40
Control Test		0.080	23:40
Air Blank		0.000	23:40
Subject Sample	# 1	0.165	23:42
Breath Volume		1.484	23:42
Air Blank		0.000	23:43
Air Blank		0.000	23:44
Subject Sample	# 2	0.162	23:47
Breath Volume		2.058	23:47
Air Blank		0.000	23:48
Control Test		0.079	23:48
Air Blank		0.000	23:48
Diagnostics Check		OK	23:48

Data Download Date/Time: 12/26/2019 14:58

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,ROBERT,E
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HAMILTON,ROBERT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: GALLEGO

DATE/TIME OF TEST: 12/07/2019 01:34:14

Diagnostics Check		OK	01:25	
Air Blank		0.000	01:26	
Control Test		0.080	01:26	
Air Blank		0.000	01:26	
Subject Sample	# 1		01:30	No Sample Provided
Breath Volume		0.000	01:30	
Air Blank		0.000	01:30	
Air Blank		0.000	01:32	
Subject Sample	# 2		01:32	Subject Test Refused
Breath Volume		0.000	01:32	
Air Blank		0.000	01:33	
Control Test		0.080	01:33	
Air Blank		0.000	01:34	
Diagnostics Check		OK	01:34	

Data Download Date/Time: 12/26/2019 14:58

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAUGHERTY,KASIE,L
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: CARTER,JACOB
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 11/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: TANNER

DATE/TIME OF TEST: 12/07/2019 23:03:16

Diagnostics Check		OK	22:58	
Air Blank		0.000	22:59	
Control Test		0.081	22:59	
Air Blank		0.000	23:00	
Subject Sample	# 1		23:01	Subject Test Refused
Breath Volume		0.000	23:01	
Air Blank		0.000	23:02	
Control Test		0.081	23:02	
Air Blank		0.000	23:03	
Diagnostics Check		OK	23:03	

Data Download Date/Time: 12/26/2019 14:58

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,ROBERT,E
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HAMPTON,JAMES
ARRESTING OFFICER AGENCY: Lake Wales PD
LAST AGENCY INSPECTION: 11/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: CUTTING III

DATE/TIME OF TEST: 12/08/2019 01:43:33

Diagnostics Check		OK	01:35	
Air Blank		0.000	01:35	
Control Test		0.080	01:35	
Air Blank		0.000	01:36	
Subject Sample	# 1		01:39	No Sample Provided
Breath Volume		0.000	01:39	
Air Blank		0.000	01:39	
Air Blank		0.000	01:41	
Subject Sample	# 2		01:42	Subject Test Refused
Breath Volume		0.000	01:42	
Air Blank		0.000	01:42	
Control Test		0.080	01:42	
Air Blank		0.000	01:43	
Diagnostics Check		OK	01:43	

Data Download Date/Time: 12/26/2019 14:58

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELCHER,ALEX,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LONDON,JOHN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: CONNELL

DATE/TIME OF TEST: 12/10/2019 19:44:27

Diagnostics Check		OK	19:35
Air Blank		0.000	19:35
Control Test		0.080	19:36
Air Blank		0.000	19:36
Subject Sample	# 1	0.234	19:39
Breath Volume		1.660	19:39
Air Blank		0.000	19:40
Air Blank		0.000	19:42
Subject Sample	# 2	0.235	19:42
Breath Volume		1.656	19:42
Air Blank		0.000	19:43
Control Test		0.079	19:43
Air Blank		0.000	19:44
Diagnostics Check		OK	19:44

Data Download Date/Time: 12/26/2019 14:58

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,ROBERT,E
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HAMILTON,ROBERT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:46

SUBJECT'S LAST NAME: AGUIRRE

DATE/TIME OF TEST: 12/12/2019 19:19:14

Diagnostics Check		OK	19:10	
Air Blank		0.000	19:11	
Control Test		0.079	19:11	
Air Blank		0.000	19:12	
Subject Sample	# 1		19:15	No Sample Provided
Breath Volume		0.000	19:15	
Air Blank		0.000	19:15	
Air Blank		0.000	19:17	
Subject Sample	# 2		19:17	Subject Test Refused
Breath Volume		0.000	19:17	
Air Blank		0.000	19:18	
Control Test		0.080	19:18	
Air Blank		0.000	19:19	
Diagnostics Check		OK	19:19	

Data Download Date/Time: 12/26/2019 14:58

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONCION,CHRISTIAN,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: MONCION,CHRISTIAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: DAIBES

DATE/TIME OF TEST: 12/13/2019 00:28:31

Diagnostics Check		OK	00:17	
Air Blank		0.000	00:17	
Control Test		0.079	00:17	
Air Blank		0.000	00:18	
Subject Sample	# 1		00:21	No Sample Provided
Breath Volume		0.000	00:21	
Air Blank		0.000	00:22	
Air Blank		0.000	00:24	
Subject Sample	# 2		00:27	No Sample Provided
Breath Volume		0.000	00:27	
Air Blank		0.000	00:27	
Control Test		0.080	00:27	
Air Blank		0.000	00:28	
Diagnostics Check		OK	00:28	

Data Download Date/Time: 12/26/2019 14:58

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COSTALES,ALEXANDRO,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: COSTALES,ALEXANDRO
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: SUAREZ

DATE/TIME OF TEST: 12/13/2019 02:17:44

Diagnostics Check		OK	02:08
Air Blank		0.000	02:09
Control Test		0.081	02:09
Air Blank		0.000	02:10
Subject Sample	# 1	0.145	02:12
Breath Volume		1.207	02:12
Air Blank		0.000	02:12
Air Blank		0.000	02:14
Subject Sample	# 2	0.147	02:16
Breath Volume		2.167	02:16
Air Blank		0.000	02:16
Control Test		0.080	02:17
Air Blank		0.000	02:17
Diagnostics Check		OK	02:17

Data Download Date/Time: 12/26/2019 14:58

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERKINS,RICHARD,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LEDFORD,ANN
ARRESTING OFFICER AGENCY: Bartow PD
LAST AGENCY INSPECTION: 11/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: FRANCOIS

DATE/TIME OF TEST: 12/13/2019 22:56:41

Diagnostics Check		OK	22:44	
Air Blank		0.000	22:45	
Control Test		0.080	22:45	
Air Blank		0.000	22:46	
Subject Sample	# 1	0.058	22:49	Volume Not Met
Breath Volume		0.289	22:49	
Air Blank		0.000	22:49	
Air Blank		0.000	22:51	
Subject Sample	# 2	0.062	22:52	
Breath Volume		1.734	22:52	
Air Blank		0.000	22:52	
Air Blank		0.000	22:54	
Subject Sample	# 3	0.061	22:55	
Breath Volume		1.902	22:55	
Air Blank		0.000	22:55	
Control Test		0.078	22:56	
Air Blank		0.000	22:56	
Diagnostics Check		OK	22:56	

Data Download Date/Time: 12/26/2019 14:58

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,MS,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DAVIS,MS
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: GARDNER

DATE/TIME OF TEST: 12/14/2019 04:43:23

Diagnostics Check		OK	04:37
Air Blank		0.000	04:37
Control Test		0.081	04:38
Air Blank		0.000	04:38
Subject Sample	# 1	0.189	04:38
Breath Volume		1.910	04:38
Air Blank		0.000	04:39
Air Blank		0.000	04:41
Subject Sample	# 2	0.182	04:41
Breath Volume		1.703	04:41
Air Blank		0.000	04:42
Control Test		0.080	04:42
Air Blank		0.000	04:43
Diagnostics Check		OK	04:43

Data Download Date/Time: 12/26/2019 14:58

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELCHER,ALEX,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: PROPST,JOHN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: AYERSMAN

DATE/TIME OF TEST: 12/15/2019 00:06:43

Diagnostics Check		OK	23:58
Air Blank		0.000	23:59
Control Test		0.081	23:59
Air Blank		0.000	23:59
Subject Sample	# 1	0.162	00:01
Breath Volume		1.597	00:01
Air Blank		0.000	00:02
Air Blank		0.000	00:04
Subject Sample	# 2	0.154	00:05
Breath Volume		2.285	00:05
Air Blank		0.000	00:05
Control Test		0.080	00:06
Air Blank		0.000	00:06
Diagnostics Check		OK	00:06

Data Download Date/Time: 12/26/2019 14:58

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERKINS,RICHARD,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GRANTHAM,CODY
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 11/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: DORLEANS

DATE/TIME OF TEST: 12/15/2019 04:07:45

Diagnostics Check		OK	04:01
Air Blank		0.000	04:01
Control Test		0.081	04:02
Air Blank		0.000	04:02
Subject Sample	# 1	0.123	04:03
Breath Volume		1.992	04:03
Air Blank		0.000	04:03
Air Blank		0.000	04:05
Subject Sample	# 2	0.115	04:06
Breath Volume		1.730	04:06
Air Blank		0.000	04:06
Control Test		0.079	04:07
Air Blank		0.000	04:07
Diagnostics Check		OK	04:07

Data Download Date/Time: 12/26/2019 14:58

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELCHER,ALEX,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 12/15/2019 22:43:11

Diagnostics Check		OK	22:35
Air Blank		0.000	22:35
Control Test		0.081	22:35
Air Blank		0.000	22:36
Subject Sample	# 1	0.243	22:38
Breath Volume		1.625	22:38
Air Blank		0.000	22:39
Air Blank		0.000	22:41
Subject Sample	# 2	0.237	22:41
Breath Volume		1.457	22:41
Air Blank		0.000	22:42
Control Test		0.079	22:42
Air Blank		0.000	22:43
Diagnostics Check		OK	22:43

Data Download Date/Time: 12/26/2019 14:58

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUCAS,NATHANIEL,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: PAPINO,ELIZABETH
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 14:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:59

SUBJECT'S LAST NAME: DUCASSE

DATE/TIME OF TEST: 12/19/2019 16:48:56

Diagnostics Check		OK	16:46	
Air Blank		0.000	16:46	
Control Test		0.080	16:46	
Air Blank		0.000	16:47	
Subject Sample	# 1		16:47	Subject Test Refused
Breath Volume		0.000	16:47	
Air Blank		0.000	16:48	
Control Test		0.080	16:48	
Air Blank		0.000	16:48	
Diagnostics Check		OK	16:48	

Data Download Date/Time: 12/26/2019 14:58

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JOHNSTON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:49

SUBJECT'S LAST NAME: UMESH

DATE/TIME OF TEST: 12/20/2019 21:26:58

Diagnostics Check		OK	21:20	
Air Blank		0.000	21:20	
Control Test		0.080	21:21	
Air Blank		0.000	21:21	
Subject Sample	# 1	0.245	21:22	
Breath Volume		1.683	21:22	
Air Blank		0.000	21:22	
Air Blank		0.000	21:24	
Subject Sample	# 2	0.232	21:25	
Breath Volume		1.390	21:25	
Air Blank		0.000	21:26	
Control Test		0.056	21:26	Control Outside Tolerance
Air Blank		0.000	21:26	

Data Download Date/Time: 12/26/2019 14:58

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAVRANEK,ADAM,J
BREATH TEST OPERATOR AGENCY: Auburndale PD
ARRESTING OFFICER: HAVRANEK,ADAM
ARRESTING OFFICER AGENCY: Auburndale PD
LAST AGENCY INSPECTION: 11/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: DUKE

DATE/TIME OF TEST: 12/20/2019 23:27:19

Diagnostics Check		OK	23:21	
Air Blank		0.000	23:21	
Control Test		0.080	23:21	
Air Blank		0.000	23:22	
Subject Sample	# 1	0.201	23:22	
Breath Volume		2.019	23:22	
Air Blank		0.000	23:23	
Air Blank		0.000	23:25	
Subject Sample	# 2	0.191	23:25	
Breath Volume		2.519	23:25	
Air Blank		0.000	23:26	
Control Test		0.070	23:26	Control Outside Tolerance
Air Blank		0.000	23:27	

Data Download Date/Time: 12/26/2019 14:58

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ-DELGADO,SHERLY,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: MATHIESON,TYSON
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 11/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: MATTUCCILLI

DATE/TIME OF TEST: 12/21/2019 03:23:31

Diagnostics Check		OK	03:16
Air Blank		0.000	03:16
Control Test		0.080	03:17
Air Blank		0.000	03:17
Subject Sample	# 1	0.205	03:18
Breath Volume		3.250	03:18
Air Blank		0.000	03:19
Air Blank		0.000	03:20
Subject Sample	# 2	0.193	03:21
Breath Volume		1.351	03:21
Air Blank		0.000	03:22
Control Test		0.080	03:22
Air Blank		0.000	03:23
Diagnostics Check		OK	03:23

Data Download Date/Time: 12/26/2019 14:58

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JOHNSTON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:31

SUBJECT'S LAST NAME: ATKINSON

DATE/TIME OF TEST: 12/21/2019 19:03:54

Diagnostics Check		OK	18:57
Air Blank		0.000	18:58
Control Test		0.080	18:58
Air Blank		0.000	18:59
Subject Sample	# 1	0.031	18:59
Breath Volume		3.593	18:59
Air Blank		0.000	19:00
Air Blank		0.000	19:01
Subject Sample	# 2	0.031	19:02
Breath Volume		4.046	19:02
Air Blank		0.000	19:02
Control Test		0.080	19:03
Air Blank		0.000	19:03
Diagnostics Check		OK	19:03

Data Download Date/Time: 12/26/2019 14:59

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COSTALES,ALEXANDRO,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: COSTALES,ALEXANDRO
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: POWELL

DATE/TIME OF TEST: 12/22/2019 01:33:17

Diagnostics Check		OK	01:21	
Air Blank		0.000	01:22	
Control Test		0.080	01:22	
Air Blank		0.000	01:22	
Subject Sample	# 1	0.180	01:26	Volume Not Met
Breath Volume		0.050	01:26	
Air Blank		0.000	01:26	
Air Blank		0.000	01:28	
Subject Sample	# 2	0.178	01:31	Volume Not Met
Breath Volume		0.015	01:31	
Air Blank		0.000	01:32	
Control Test		0.078	01:32	
Air Blank		0.000	01:33	
Diagnostics Check		OK	01:33	

Data Download Date/Time: 12/26/2019 14:59

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COSTALES,ALEXANDRO,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: COSTALES,ALEXANDRO
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: POWELL

DATE/TIME OF TEST: 12/22/2019 01:46:43

Diagnostics Check		OK	01:35	
Air Blank		0.000	01:35	
Control Test		0.079	01:35	
Air Blank		0.000	01:36	
Subject Sample	# 1	0.174	01:39	Volume Not Met
Breath Volume		0.316	01:39	
Air Blank		0.000	01:40	
Air Blank		0.000	01:42	
Subject Sample	# 2	0.178	01:45	Volume Not Met
Breath Volume		0.046	01:45	
Air Blank		0.000	01:45	
Control Test		0.079	01:46	
Air Blank		0.000	01:46	
Diagnostics Check		OK	01:46	

Data Download Date/Time: 12/26/2019 14:59

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUGAN,LAKEBRA,S
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SIDLER,JAMIE
ARRESTING OFFICER AGENCY: Bartow PD
LAST AGENCY INSPECTION: 12/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 12/25/2019 00:12:10

Diagnostics Check		OK	00:03
Air Blank		0.000	00:03
Control Test		0.080	00:03
Air Blank		0.000	00:04
Subject Sample	# 1	0.258	00:07
Breath Volume		4.003	00:07
Air Blank		0.000	00:08
Air Blank		0.000	00:10
Subject Sample	# 2	0.247	00:10
Breath Volume		1.609	00:10
Air Blank		0.000	00:11
Control Test		0.079	00:11
Air Blank		0.000	00:12
Diagnostics Check		OK	00:12

Data Download Date/Time: 12/26/2019 14:59

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONCION,CHRISTIAN,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: MONCION,CHRISTIAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: DIAZ RONDON

DATE/TIME OF TEST: 12/25/2019 01:03:15

Diagnostics Check		OK	00:54
Air Blank		0.000	00:54
Control Test		0.080	00:55
Air Blank		0.000	00:55
Subject Sample	# 1	0.220	00:58
Breath Volume		1.144	00:58
Air Blank		0.000	00:59
Air Blank		0.000	01:00
Subject Sample	# 2	0.209	01:01
Breath Volume		1.183	01:01
Air Blank		0.000	01:02
Control Test		0.078	01:02
Air Blank		0.000	01:03
Diagnostics Check		OK	01:03

Data Download Date/Time: 12/26/2019 14:59

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES,STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SYKES,STEVEN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:28

SUBJECT'S LAST NAME: NIPPER

DATE/TIME OF TEST: 12/25/2019 21:54:57

Diagnostics Check		OK	21:48
Air Blank		0.000	21:48
Control Test		0.080	21:49
Air Blank		0.000	21:49
Subject Sample	# 1	0.274	21:50
Breath Volume		1.410	21:50
Air Blank		0.000	21:50
Air Blank		0.000	21:52
Subject Sample	# 2	0.286	21:53
Breath Volume		1.378	21:53
Air Blank		0.000	21:54
Control Test		0.080	21:54
Air Blank		0.000	21:54
Diagnostics Check		OK	21:54

Data Download Date/Time: 12/26/2019 14:59

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES,STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JOHNSTON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: LOPEZ GOMEZ

DATE/TIME OF TEST: 12/27/2019 00:03:45

Diagnostics Check		OK	23:50	
Air Blank		0.000	23:50	
Control Test		0.080	23:51	
Air Blank		0.000	23:51	
Subject Sample	# 2		23:54	No Sample Provided
Breath Volume		0.000	23:54	
Air Blank		0.000	23:55	
Air Blank		0.000	23:57	
Subject Sample	# 3	0.180	23:58	
Breath Volume		2.300	23:58	
Air Blank		0.000	23:59	
Air Blank		0.000	00:01	
Subject Sample	# 1	0.185	00:02	
Breath Volume		1.703	00:02	
Air Blank		0.000	00:02	
Control Test		0.080	00:03	
Air Blank		0.000	00:03	
Diagnostics Check		OK	00:03	

Data Download Date/Time: 01/27/2020 09:04

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,DALE,E
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ROBERTS,DALE
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:33

SUBJECT'S LAST NAME: CONTRERAS

DATE/TIME OF TEST: 12/29/2019 05:57:48

Diagnostics Check		OK	05:54	
Air Blank		0.000	05:54	
Control Test		0.081	05:55	
Air Blank		0.000	05:55	
Subject Sample	# 1		05:56	Subject Test Refused
Breath Volume		0.000	05:56	
Air Blank		0.000	05:56	
Control Test		0.081	05:57	
Air Blank		0.000	05:57	
Diagnostics Check		OK	05:57	

Data Download Date/Time: 01/27/2020 09:04

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JOHNSTON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 12/30/2019 22:29:28

Diagnostics Check		OK	22:23
Air Blank		0.000	22:23
Control Test		0.080	22:23
Air Blank		0.000	22:24
Subject Sample	# 1	0.246	22:24
Breath Volume		3.011	22:24
Air Blank		0.000	22:25
Air Blank		0.000	22:27
Subject Sample	# 2	0.256	22:27
Breath Volume		3.140	22:27
Air Blank		0.000	22:28
Control Test		0.080	22:28
Air Blank		0.000	22:29
Diagnostics Check		OK	22:29

Data Download Date/Time: 01/27/2020 09:04

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIGGINBOTHAM,JACK,C
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HIGGINBOTHAM,JACK
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: PASTORKY

DATE/TIME OF TEST: 12/30/2019 22:48:06

Diagnostics Check		OK	22:41
Air Blank		0.000	22:41
Control Test		0.080	22:42
Air Blank		0.000	22:42
Subject Sample	# 1	0.000	22:43
Breath Volume		2.167	22:43
Air Blank		0.000	22:43
Air Blank		0.000	22:45
Subject Sample	# 2	0.000	22:46
Breath Volume		3.070	22:46
Air Blank		0.000	22:47
Control Test		0.080	22:47
Air Blank		0.000	22:48
Diagnostics Check		OK	22:48

Data Download Date/Time: 01/27/2020 09:04

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,ROBERT,E
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HAMILTON,ROBERT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: CRUMMEY

DATE/TIME OF TEST: 12/31/2019 00:58:14

Diagnostics Check		OK	00:51
Air Blank		0.000	00:52
Control Test		0.080	00:52
Air Blank		0.000	00:53
Subject Sample	# 1	0.172	00:53
Breath Volume		3.062	00:53
Air Blank		0.000	00:54
Air Blank		0.000	00:56
Subject Sample	# 2	0.171	00:56
Breath Volume		3.171	00:56
Air Blank		0.000	00:57
Control Test		0.079	00:57
Air Blank		0.000	00:58
Diagnostics Check		OK	00:58

Data Download Date/Time: 01/27/2020 09:04

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,ROBERT,E
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: PICHE,RUSSELL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: CHATMAN

DATE/TIME OF TEST: 12/31/2019 23:00:29

Diagnostics Check		OK	22:53
Air Blank		0.000	22:54
Control Test		0.080	22:54
Air Blank		0.000	22:54
Subject Sample	# 1	0.140	22:56
Breath Volume		3.500	22:56
Air Blank		0.000	22:56
Air Blank		0.000	22:58
Subject Sample	# 2	0.141	22:58
Breath Volume		3.667	22:58
Air Blank		0.000	22:59
Control Test		0.079	22:59
Air Blank		0.000	23:00
Diagnostics Check		OK	23:00

Data Download Date/Time: 01/27/2020 09:04

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELCHER,ALEX,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ALLRED,COBY
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: JEETOO

DATE/TIME OF TEST: 12/31/2019 23:18:45

Diagnostics Check		OK	23:05	
Air Blank		0.000	23:06	
Control Test		0.079	23:06	
Air Blank		0.000	23:06	
Subject Sample	# 1	0.168	23:09	Volume Not Met
Breath Volume		0.125	23:09	
Air Blank		0.000	23:10	
Air Blank		0.000	23:12	
Subject Sample	# 2	0.172	23:13	
Breath Volume		1.136	23:13	
Air Blank		0.000	23:14	
Air Blank		0.000	23:15	
Subject Sample	# 3	0.189	23:17	
Breath Volume		1.886	23:17	
Air Blank		0.000	23:17	
Control Test		0.079	23:18	
Air Blank		0.000	23:18	
Diagnostics Check		OK	23:18	

Data Download Date/Time: 01/27/2020 09:04

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JOHNSTON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: JUARES

DATE/TIME OF TEST: 12/31/2019 23:51:22

Diagnostics Check		OK	23:39	
Air Blank		0.000	23:40	
Control Test		0.079	23:40	
Air Blank		0.000	23:41	
Subject Sample	# 1		23:44	No Sample Provided
Breath Volume		0.000	23:44	
Air Blank		0.000	23:44	
Air Blank		0.000	23:46	
Subject Sample	# 2		23:50	No Sample Provided
Breath Volume		0.000	23:50	
Air Blank		0.000	23:50	
Control Test		0.079	23:50	
Air Blank		0.000	23:51	
Diagnostics Check		OK	23:51	

Data Download Date/Time: 01/27/2020 09:04

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

INSTRUMENT REGISTERED TO: DAVIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,SHELIA,R
BREATH TEST OPERATOR AGENCY: Davie PD
ARRESTING OFFICER: PUENTES,YENDI
ARRESTING OFFICER AGENCY: Davie PD
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: TALAVERA RUIZ

DATE/TIME OF TEST: 12/02/2019 02:45:37

Diagnostics Check		OK	02:38
Air Blank		0.000	02:38
Control Test		0.081	02:38
Air Blank		0.000	02:39
Subject Sample	# 1	0.254	02:40
Breath Volume		1.949	02:40
Air Blank		0.000	02:41
Air Blank		0.000	02:42
Subject Sample	# 2	0.261	02:44
Breath Volume		2.535	02:44
Air Blank		0.000	02:44
Control Test		0.079	02:45
Air Blank		0.000	02:45
Diagnostics Check		OK	02:45

Data Download Date/Time: 01/27/2020 11:41

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

INSTRUMENT REGISTERED TO: DAVIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAHAM,HUGH,E
BREATH TEST OPERATOR AGENCY: Davie PD
ARRESTING OFFICER: CORREA,MATTHEW
ARRESTING OFFICER AGENCY: Davie PD
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: RASHID

DATE/TIME OF TEST: 12/09/2019 01:54:30

Diagnostics Check		OK	01:47
Air Blank		0.000	01:47
Control Test		0.081	01:47
Air Blank		0.000	01:48
Subject Sample	# 1	0.126	01:49
Breath Volume		1.136	01:49
Air Blank		0.000	01:50
Air Blank		0.000	01:51
Subject Sample	# 2	0.122	01:52
Breath Volume		1.132	01:52
Air Blank		0.000	01:53
Control Test		0.080	01:53
Air Blank		0.000	01:54
Diagnostics Check		OK	01:54

Data Download Date/Time: 01/27/2020 11:41

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

INSTRUMENT REGISTERED TO: DAVIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAINES,DAVID,K
BREATH TEST OPERATOR AGENCY: Davie PD
ARRESTING OFFICER: YOUNG,JOSHUA
ARRESTING OFFICER AGENCY: Davie PD
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: PINTO

DATE/TIME OF TEST: 12/13/2019 03:27:27

Diagnostics Check		OK	03:20
Air Blank		0.000	03:21
Control Test		0.081	03:21
Air Blank		0.000	03:22
Subject Sample	# 1	0.174	03:22
Breath Volume		3.285	03:22
Air Blank		0.000	03:23
Air Blank		0.000	03:25
Subject Sample	# 2	0.163	03:25
Breath Volume		3.222	03:25
Air Blank		0.000	03:26
Control Test		0.081	03:26
Air Blank		0.000	03:27
Diagnostics Check		OK	03:27

Data Download Date/Time: 01/27/2020 11:41

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

INSTRUMENT REGISTERED TO: DAVIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,SHELIA,R
BREATH TEST OPERATOR AGENCY: Davie PD
ARRESTING OFFICER: HACKETT,JULIA
ARRESTING OFFICER AGENCY: Davie PD
LAST AGENCY INSPECTION: 11/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 07:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:33

SUBJECT'S LAST NAME: OCHOA SUAREZ

DATE/TIME OF TEST: 12/14/2019 08:01:35

Diagnostics Check		OK	07:55
Air Blank		0.000	07:55
Control Test		0.080	07:56
Air Blank		0.000	07:56
Subject Sample	# 1	0.169	07:57
Breath Volume		2.734	07:57
Air Blank		0.000	07:57
Air Blank		0.000	07:59
Subject Sample	# 2	0.178	08:00
Breath Volume		2.449	08:00
Air Blank		0.000	08:00
Control Test		0.081	08:00
Air Blank		0.000	08:01
Diagnostics Check		OK	08:01

Data Download Date/Time: 01/27/2020 11:41

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

INSTRUMENT REGISTERED TO: GULFPORT PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLAGUE,ERIC,A
BREATH TEST OPERATOR AGENCY: Gulfport PD
ARRESTING OFFICER: CLAGUE,ERIC
ARRESTING OFFICER AGENCY: Gulfport PD
LAST AGENCY INSPECTION: 11/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: DUGMORE

DATE/TIME OF TEST: 12/15/2019 03:02:13

Diagnostics Check		OK	02:59	
Air Blank		0.000	02:59	
Control Test		0.079	03:00	
Air Blank		0.000	03:00	
Subject Sample	# 1		03:00	Subject Test Refused
Breath Volume		0.000	03:00	
Air Blank		0.000	03:01	
Control Test		0.080	03:01	
Air Blank		0.000	03:02	
Diagnostics Check		OK	03:02	

Data Download Date/Time: 01/07/2020 14:48

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

INSTRUMENT REGISTERED TO: GULFPORT PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,TIMOTHY,R
BREATH TEST OPERATOR AGENCY: Gulfport PD
ARRESTING OFFICER: CARTER,TIMOTHY
ARRESTING OFFICER AGENCY: Gulfport PD
LAST AGENCY INSPECTION: 11/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: MAZZOLA

DATE/TIME OF TEST: 12/28/2019 22:45:22

Diagnostics Check		OK	22:36	
Air Blank		0.000	22:37	
Control Test		0.079	22:37	
Air Blank		0.000	22:38	
Subject Sample	# 1		22:41	No Sample Provided
Breath Volume		0.000	22:41	
Air Blank		0.000	22:41	
Air Blank		0.000	22:43	
Subject Sample	# 2		22:44	Subject Test Refused
Breath Volume		0.000	22:44	
Air Blank		0.000	22:44	
Control Test		0.079	22:44	
Air Blank		0.000	22:45	
Diagnostics Check		OK	22:45	

Data Download Date/Time: 01/07/2020 14:48

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

INSTRUMENT REGISTERED TO: GULFPORT PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRIEST,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Gulfport PD
ARRESTING OFFICER: PRIEST,CHRISTOPHER
ARRESTING OFFICER AGENCY: Gulfport PD
LAST AGENCY INSPECTION: 11/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 07:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:20

SUBJECT'S LAST NAME: RUSSELL

DATE/TIME OF TEST: 12/29/2019 07:52:18

Diagnostics Check		OK	07:40	
Air Blank		0.000	07:41	
Control Test		0.080	07:41	
Air Blank		0.000	07:42	
Subject Sample	# 1		07:45	No Sample Provided
Breath Volume		0.000	07:45	
Air Blank		0.000	07:45	
Air Blank		0.000	07:47	
Subject Sample	# 2		07:50	No Sample Provided
Breath Volume		0.000	07:50	
Air Blank		0.000	07:51	
Control Test		0.080	07:51	
Air Blank		0.000	07:52	
Diagnostics Check		OK	07:52	

Data Download Date/Time: 01/07/2020 14:48

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

INSTRUMENT REGISTERED TO: HOLLYWOOD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,COREY,
BREATH TEST OPERATOR AGENCY: Hollywood PD
ARRESTING OFFICER: WILLIAMS,COREY
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 11/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 12:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:49

SUBJECT'S LAST NAME: CASHDOLLAR

DATE/TIME OF TEST: 12/02/2019 13:19:50

Diagnostics Check		OK	13:13
Air Blank		0.000	13:13
Control Test		0.079	13:14
Air Blank		0.000	13:14
Subject Sample	# 1	0.311	13:15
Breath Volume		2.976	13:15
Air Blank		0.000	13:15
Air Blank		0.000	13:17
Subject Sample	# 2	0.306	13:18
Breath Volume		2.816	13:18
Air Blank		0.000	13:18
Control Test		0.078	13:19
Air Blank		0.000	13:19
Diagnostics Check		OK	13:19

Data Download Date/Time: 12/26/2019 11:37

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Port Richey PD
ARRESTING OFFICER: HAZIME,A
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 11/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: KNOWLES

DATE/TIME OF TEST: 12/05/2019 04:01:22

Diagnostics Check		OK	03:54
Air Blank		0.000	03:54
Control Test		0.078	03:55
Air Blank		0.000	03:55
Subject Sample	# 1	0.023	03:56
Breath Volume		2.324	03:56
Air Blank		0.000	03:57
Air Blank		0.000	03:59
Subject Sample	# 2	0.022	03:59
Breath Volume		2.367	03:59
Air Blank		0.000	04:00
Control Test		0.077	04:00
Air Blank		0.000	04:01
Diagnostics Check		OK	04:01

Data Download Date/Time: 12/26/2019 22:56

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAZIME,OFC. A.,
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: HAMILTON,OFC B.
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 11/24/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: CLUKEY

DATE/TIME OF TEST: 12/14/2019 00:58:33

Diagnostics Check		OK	00:52
Air Blank		0.000	00:52
Control Test		0.078	00:52
Air Blank		0.000	00:53
Subject Sample	# 1	0.170	00:54
Breath Volume		2.101	00:54
Air Blank		0.000	00:54
Air Blank		0.000	00:56
Subject Sample	# 2	0.177	00:57
Breath Volume		1.511	00:57
Air Blank		0.000	00:57
Control Test		0.078	00:57
Air Blank		0.000	00:58
Diagnostics Check		OK	00:58

Data Download Date/Time: 12/26/2019 22:56

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAZIME,OFC. A.,
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: KASPERITIS,DEP A.
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 11/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 05:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:08

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 12/15/2019 06:35:10

Diagnostics Check		OK	06:29
Air Blank		0.000	06:29
Control Test		0.078	06:29
Air Blank		0.000	06:30
Subject Sample	# 1	0.151	06:30
Breath Volume		1.878	06:30
Air Blank		0.000	06:31
Air Blank		0.000	06:33
Subject Sample	# 2	0.153	06:33
Breath Volume		2.011	06:33
Air Blank		0.000	06:34
Control Test		0.078	06:34
Air Blank		0.000	06:35
Diagnostics Check		OK	06:35

Data Download Date/Time: 12/26/2019 22:56

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Port Richey PD
ARRESTING OFFICER: BABCOCK,EDWARD
ARRESTING OFFICER AGENCY: Port Richey PD
LAST AGENCY INSPECTION: 11/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: TOMLINSON

DATE/TIME OF TEST: 12/16/2019 02:30:41

Diagnostics Check		OK	02:21
Air Blank		0.000	02:22
Control Test		0.077	02:22
Air Blank		0.000	02:23
Subject Sample	# 1	0.205	02:25
Breath Volume		3.105	02:25
Air Blank		0.000	02:25
Air Blank		0.000	02:27
Subject Sample	# 2	0.205	02:29
Breath Volume		4.710	02:29
Air Blank		0.000	02:29
Control Test		0.077	02:30
Air Blank		0.000	02:30
Diagnostics Check		OK	02:30

Data Download Date/Time: 12/26/2019 22:56

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAZIME,OFC. A.,
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: HAZIME,OFC. A.
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 11/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 12/24/2019 00:41:49

Diagnostics Check		OK	00:35
Air Blank		0.000	00:35
Control Test		0.077	00:36
Air Blank		0.000	00:36
Subject Sample	# 1	0.229	00:36
Breath Volume		1.863	00:36
Air Blank		0.000	00:37
Air Blank		0.000	00:39
Subject Sample	# 2	0.225	00:40
Breath Volume		1.925	00:40
Air Blank		0.000	00:40
Control Test		0.078	00:41
Air Blank		0.000	00:41
Diagnostics Check		OK	00:41

Data Download Date/Time: 12/26/2019 22:56

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCONNELL,J,W
BREATH TEST OPERATOR AGENCY: Port Richey PD
ARRESTING OFFICER: MCCONNELL,J
ARRESTING OFFICER AGENCY: Port Richey PD
LAST AGENCY INSPECTION: 11/24/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: MERIDITH

DATE/TIME OF TEST: 12/26/2019 00:48:04

Diagnostics Check		OK	00:41
Air Blank		0.000	00:42
Control Test		0.077	00:42
Air Blank		0.000	00:43
Subject Sample	# 1	0.310	00:43
Breath Volume		1.296	00:43
Air Blank		0.000	00:44
Air Blank		0.000	00:45
Subject Sample	# 2	0.310	00:46
Breath Volume		1.652	00:46
Air Blank		0.000	00:47
Control Test		0.077	00:47
Air Blank		0.000	00:47
Diagnostics Check		OK	00:48

Data Download Date/Time: 12/26/2019 22:56

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCONNELL,J,W
BREATH TEST OPERATOR AGENCY: Port Richey PD
ARRESTING OFFICER: MCCONNELL,J
ARRESTING OFFICER AGENCY: Port Richey PD
LAST AGENCY INSPECTION: 12/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 12/31/2019 23:53:10

Diagnostics Check		OK	23:41	
Air Blank		0.000	23:41	
Control Test		0.077	23:42	
Air Blank		0.000	23:42	
Subject Sample	# 1	0.163	23:45	Volume Not Met
Breath Volume		0.117	23:45	
Air Blank		0.000	23:46	
Air Blank		0.000	23:48	
Subject Sample	# 2	0.132	23:51	Volume Not Met
Breath Volume		0.132	23:51	
Air Blank		0.000	23:52	
Control Test		0.076	23:52	
Air Blank		0.000	23:53	
Diagnostics Check		OK	23:53	

Data Download Date/Time: 01/28/2020 22:17

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: CAPPER

DATE/TIME OF TEST: 12/01/2019 01:03:46

Diagnostics Check		OK	01:00	
Air Blank		0.000	01:01	
Control Test		0.082	01:01	
Air Blank		0.000	01:02	
Subject Sample	# 1		01:02	Subject Test Refused
Breath Volume		0.000	01:02	
Air Blank		0.000	01:02	
Control Test		0.081	01:03	
Air Blank		0.000	01:03	
Diagnostics Check		OK	01:03	

Data Download Date/Time: 12/21/2019 19:27

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MASON,G
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: RIGGINS

DATE/TIME OF TEST: 12/01/2019 02:09:21

Diagnostics Check		OK	01:59
Air Blank		0.000	01:59
Control Test		0.082	01:59
Air Blank		0.000	02:00
Subject Sample	# 1	0.216	02:01
Breath Volume		2.164	02:01
Air Blank		0.000	02:01
Air Blank		0.000	02:03
Subject Sample	# 2	0.180	02:04
Breath Volume		1.496	02:04
Air Blank		0.000	02:05
Air Blank		0.000	02:06
Subject Sample	# 3	0.200	02:07
Breath Volume		2.128	02:07
Air Blank		0.000	02:08
Control Test		0.081	02:08
Air Blank		0.000	02:09
Diagnostics Check		OK	02:09

Data Download Date/Time: 12/21/2019 19:27

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SPATZ,J
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: MCCOMAS

DATE/TIME OF TEST: 12/01/2019 02:26:51

Diagnostics Check		OK	02:20
Air Blank		0.000	02:20
Control Test		0.081	02:21
Air Blank		0.000	02:21
Subject Sample	# 1	0.184	02:22
Breath Volume		2.050	02:22
Air Blank		0.000	02:22
Air Blank		0.000	02:24
Subject Sample	# 2	0.187	02:25
Breath Volume		2.609	02:25
Air Blank		0.000	02:25
Control Test		0.081	02:26
Air Blank		0.000	02:26
Diagnostics Check		OK	02:26

Data Download Date/Time: 12/21/2019 19:27

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PEYINGHOUSE,D
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: RABNER

DATE/TIME OF TEST: 12/01/2019 04:19:03

Diagnostics Check		OK	04:15	
Air Blank		0.000	04:16	
Control Test		0.082	04:16	
Air Blank		0.000	04:17	
Subject Sample	# 1		04:17	Subject Test Refused
Breath Volume		0.000	04:17	
Air Blank		0.000	04:18	
Control Test		0.081	04:18	
Air Blank		0.000	04:18	
Diagnostics Check		OK	04:19	

Data Download Date/Time: 12/21/2019 19:27

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICOTTILLI,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI,JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: BRESEMAN

DATE/TIME OF TEST: 12/01/2019 04:51:49

Diagnostics Check		OK	04:47	
Air Blank		0.000	04:47	
Control Test		0.082	04:47	
Air Blank		0.000	04:48	
Subject Sample	# 1		04:50	Subject Test Refused
Breath Volume		0.000	04:50	
Air Blank		0.000	04:50	
Control Test		0.081	04:51	
Air Blank		0.000	04:51	
Diagnostics Check		OK	04:51	

Data Download Date/Time: 12/21/2019 19:27

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,WILLIAM,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 06:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:43

SUBJECT'S LAST NAME: COGGESHALL

DATE/TIME OF TEST: 12/01/2019 08:12:34

Diagnostics Check		OK	08:06
Air Blank		0.000	08:06
Control Test		0.082	08:07
Air Blank		0.000	08:07
Subject Sample	# 1	0.000	08:08
Breath Volume		2.156	08:08
Air Blank		0.000	08:08
Air Blank		0.000	08:10
Subject Sample	# 2	0.000	08:11
Breath Volume		1.777	08:11
Air Blank		0.000	08:11
Control Test		0.082	08:11
Air Blank		0.000	08:12
Diagnostics Check		OK	08:12

Data Download Date/Time: 12/21/2019 19:27

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WILLIAMS,TIMMY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 11:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:05

SUBJECT'S LAST NAME: SPERLING

DATE/TIME OF TEST: 12/03/2019 12:33:50

Diagnostics Check		OK	12:25
Air Blank		0.000	12:26
Control Test		0.082	12:26
Air Blank		0.000	12:27
Subject Sample	# 1	0.000	12:28
Breath Volume		1.812	12:28
Air Blank		0.000	12:28
Air Blank		0.000	12:30
Subject Sample	# 2	0.000	12:32
Breath Volume		2.570	12:32
Air Blank		0.000	12:32
Control Test		0.082	12:33
Air Blank		0.000	12:33
Diagnostics Check		OK	12:33

Data Download Date/Time: 12/21/2019 19:27

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 18:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:43

SUBJECT'S LAST NAME: COLBERT

DATE/TIME OF TEST: 12/03/2019 19:10:22

Diagnostics Check		OK	19:03
Air Blank		0.000	19:03
Control Test		0.083	19:04
Air Blank		0.000	19:04
Subject Sample	# 1	0.152	19:05
Breath Volume		2.082	19:05
Air Blank		0.000	19:06
Air Blank		0.000	19:07
Subject Sample	# 2	0.151	19:08
Breath Volume		1.800	19:08
Air Blank		0.000	19:09
Control Test		0.081	19:09
Air Blank		0.000	19:10
Diagnostics Check		OK	19:10

Data Download Date/Time: 12/21/2019 19:27

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELSHANS,DEAN,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MARTINI,ISABELLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 19:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:31

SUBJECT'S LAST NAME: SAWAYA

DATE/TIME OF TEST: 12/03/2019 21:02:25

Diagnostics Check		OK	20:53	
Air Blank		0.000	20:54	
Control Test		0.082	20:54	
Air Blank		0.000	20:54	
Subject Sample	# 1	0.114	20:58	
Breath Volume		1.914	20:58	
Air Blank		0.000	20:58	
Air Blank		0.000	21:00	
Subject Sample	# 2		21:01	RFI Detect
Breath Volume		0.000	21:01	
Air Blank		0.000	21:01	
Control Test		0.081	21:01	
Air Blank		0.000	21:02	
Diagnostics Check		OK	21:02	

Data Download Date/Time: 12/21/2019 19:27

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELSHANS,DEAN,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MARTINI,ISABELLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 19:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:31

SUBJECT'S LAST NAME: SAWAYA

DATE/TIME OF TEST: 12/03/2019 21:11:59

Diagnostics Check		OK	21:05
Air Blank		0.000	21:05
Control Test		0.081	21:06
Air Blank		0.000	21:06
Subject Sample	# 1	0.109	21:07
Breath Volume		1.464	21:07
Air Blank		0.000	21:08
Air Blank		0.000	21:09
Subject Sample	# 2	0.109	21:10
Breath Volume		1.695	21:10
Air Blank		0.000	21:11
Control Test		0.082	21:11
Air Blank		0.000	21:11
Diagnostics Check		OK	21:11

Data Download Date/Time: 12/21/2019 19:27

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: STUKUS,BRANDON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: MORICZ

DATE/TIME OF TEST: 12/04/2019 01:21:10

Diagnostics Check		OK	01:12
Air Blank		0.000	01:13
Control Test		0.083	01:13
Air Blank		0.000	01:14
Subject Sample	# 1	0.148	01:15
Breath Volume		1.410	01:15
Air Blank		0.000	01:16
Air Blank		0.000	01:18
Subject Sample	# 2	0.131	01:19
Breath Volume		2.140	01:19
Air Blank		0.000	01:20
Control Test		0.082	01:20
Air Blank		0.000	01:21
Diagnostics Check		OK	01:21

Data Download Date/Time: 12/21/2019 19:27

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSEN,STEPHEN,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: JACOBSEN,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: BONACCI

DATE/TIME OF TEST: 12/04/2019 02:34:39

Diagnostics Check		OK	02:21	
Air Blank		0.000	02:22	
Control Test		0.080	02:22	
Air Blank		0.000	02:23	
Subject Sample	# 1	0.171	02:26	Volume Not Met
Breath Volume		0.523	02:26	
Air Blank		0.000	02:26	
Air Blank		0.000	02:28	
Subject Sample	# 2	0.216	02:29	
Breath Volume		1.757	02:29	
Air Blank		0.000	02:29	
Air Blank		0.000	02:31	
Subject Sample	# 3	0.202	02:33	
Breath Volume		1.312	02:33	
Air Blank		0.000	02:33	
Control Test		0.081	02:34	
Air Blank		0.000	02:34	
Diagnostics Check		OK	02:34	

Data Download Date/Time: 12/21/2019 19:27

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: MARCELLUS

DATE/TIME OF TEST: 12/04/2019 23:22:57

Diagnostics Check		OK	23:11	
Air Blank		0.000	23:11	
Control Test		0.082	23:11	
Air Blank		0.000	23:12	
Subject Sample	# 1	0.169	23:15	Volume Not Met
Breath Volume		0.101	23:15	
Air Blank		0.000	23:16	
Air Blank		0.000	23:18	
Subject Sample	# 2	0.213	23:21	Volume Not Met
Breath Volume		0.152	23:21	
Air Blank		0.000	23:21	
Control Test		0.080	23:22	
Air Blank		0.000	23:22	
Diagnostics Check		OK	23:22	

Data Download Date/Time: 12/21/2019 19:27

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: MARCELLUS

DATE/TIME OF TEST: 12/04/2019 23:36:48

Diagnostics Check		OK	23:25	
Air Blank		0.000	23:25	
Control Test		0.079	23:25	
Air Blank		0.000	23:26	
Subject Sample	# 1	0.206	23:29	Volume Not Met
Breath Volume		0.039	23:29	
Air Blank		0.000	23:30	
Air Blank		0.000	23:32	
Subject Sample	# 2	0.190	23:35	Volume Not Met
Breath Volume		0.101	23:35	
Air Blank		0.000	23:35	
Control Test		0.079	23:36	
Air Blank		0.000	23:36	
Diagnostics Check		OK	23:36	

Data Download Date/Time: 12/21/2019 19:27

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: MARCELLUS

DATE/TIME OF TEST: 12/04/2019 23:50:57

Diagnostics Check		OK	23:39	
Air Blank		0.000	23:39	
Control Test		0.080	23:40	
Air Blank		0.000	23:40	
Subject Sample	# 1	0.173	23:43	Volume Not Met
Breath Volume		0.316	23:43	
Air Blank		0.000	23:44	
Air Blank		0.000	23:46	
Subject Sample	# 2	0.199	23:49	Volume Not Met
Breath Volume		0.015	23:49	
Air Blank		0.000	23:50	
Control Test		0.080	23:50	
Air Blank		0.000	23:50	
Diagnostics Check		OK	23:50	

Data Download Date/Time: 12/21/2019 19:27

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 12/05/2019 00:37:06

Diagnostics Check		OK	00:31
Air Blank		0.000	00:31
Control Test		0.079	00:31
Air Blank		0.000	00:32
Subject Sample	# 1	0.166	00:32
Breath Volume		1.554	00:32
Air Blank		0.000	00:33
Air Blank		0.000	00:35
Subject Sample	# 2	0.167	00:35
Breath Volume		1.414	00:35
Air Blank		0.000	00:36
Control Test		0.080	00:36
Air Blank		0.000	00:37
Diagnostics Check		OK	00:37

Data Download Date/Time: 12/21/2019 19:27

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANWORP,JUDSON,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: VANWORP,JUDSON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 12/05/2019 01:07:27

Diagnostics Check		OK	00:59
Air Blank		0.000	00:59
Control Test		0.080	00:59
Air Blank		0.000	01:00
Subject Sample	# 1	0.279	01:02
Breath Volume		1.500	01:02
Air Blank		0.000	01:03
Air Blank		0.000	01:04
Subject Sample	# 2	0.272	01:05
Breath Volume		1.574	01:05
Air Blank		0.000	01:06
Control Test		0.078	01:06
Air Blank		0.000	01:07
Diagnostics Check		OK	01:07

Data Download Date/Time: 12/21/2019 19:27

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BEARY,F
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: BENTLEY

DATE/TIME OF TEST: 12/05/2019 02:35:55

Diagnostics Check		OK	02:29
Air Blank		0.000	02:29
Control Test		0.081	02:29
Air Blank		0.000	02:30
Subject Sample	# 1	0.157	02:31
Breath Volume		1.878	02:31
Air Blank		0.000	02:32
Air Blank		0.000	02:34
Subject Sample	# 2	0.161	02:34
Breath Volume		1.742	02:34
Air Blank		0.000	02:35
Control Test		0.080	02:35
Air Blank		0.000	02:35
Diagnostics Check		OK	02:35

Data Download Date/Time: 12/21/2019 19:27

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: WESTMORELAND

DATE/TIME OF TEST: 12/05/2019 04:02:26

Diagnostics Check		OK	03:59	
Air Blank		0.000	03:59	
Control Test		0.082	04:00	
Air Blank		0.000	04:00	
Subject Sample	# 1		04:01	Subject Test Refused
Breath Volume		0.000	04:01	
Air Blank		0.000	04:01	
Control Test		0.081	04:01	
Air Blank		0.000	04:02	
Diagnostics Check		OK	04:02	

Data Download Date/Time: 12/21/2019 19:27

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 03:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: ARNTZ

DATE/TIME OF TEST: 12/05/2019 04:47:03

Diagnostics Check		OK	04:38
Air Blank		0.000	04:39
Control Test		0.081	04:39
Air Blank		0.000	04:40
Subject Sample	# 1	0.117	04:40
Breath Volume		1.605	04:40
Air Blank		0.000	04:41
Air Blank		0.000	04:42
Subject Sample	# 2	0.113	04:45
Breath Volume		2.492	04:45
Air Blank		0.000	04:46
Control Test		0.081	04:46
Air Blank		0.000	04:46
Diagnostics Check		OK	04:47

Data Download Date/Time: 12/21/2019 19:27

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: ATKINS

DATE/TIME OF TEST: 12/06/2019 02:00:35

Diagnostics Check		OK	01:57	
Air Blank		0.000	01:57	
Control Test		0.082	01:58	
Air Blank		0.000	01:58	
Subject Sample	# 1		01:59	Subject Test Refused
Breath Volume		0.000	01:59	
Air Blank		0.000	01:59	
Control Test		0.081	02:00	
Air Blank		0.000	02:00	
Diagnostics Check		OK	02:00	

Data Download Date/Time: 12/21/2019 19:27

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 06:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:54

SUBJECT'S LAST NAME: LINGERFELT

DATE/TIME OF TEST: 12/06/2019 07:22:23

Diagnostics Check		OK	07:14
Air Blank		0.000	07:14
Control Test		0.082	07:15
Air Blank		0.000	07:15
Subject Sample	# 1	0.029	07:17
Breath Volume		1.648	07:17
Air Blank		0.000	07:18
Air Blank		0.000	07:19
Subject Sample	# 2	0.029	07:20
Breath Volume		2.246	07:20
Air Blank		0.000	07:21
Control Test		0.082	07:21
Air Blank		0.000	07:22
Diagnostics Check		OK	07:22

Data Download Date/Time: 12/21/2019 19:27

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICOTTILLI,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI,JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 18:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:47

SUBJECT'S LAST NAME: PERONIS

DATE/TIME OF TEST: 12/06/2019 19:16:01

Diagnostics Check		OK	19:07
Air Blank		0.000	19:07
Control Test		0.082	19:08
Air Blank		0.000	19:08
Subject Sample	# 1	0.048	19:10
Breath Volume		1.597	19:10
Air Blank		0.000	19:10
Air Blank		0.000	19:12
Subject Sample	# 2	0.045	19:14
Breath Volume		1.320	19:14
Air Blank		0.000	19:15
Control Test		0.082	19:15
Air Blank		0.000	19:15
Diagnostics Check		OK	19:16

Data Download Date/Time: 12/21/2019 19:27

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:43

SUBJECT'S LAST NAME: DEMA

DATE/TIME OF TEST: 12/06/2019 21:10:36

Diagnostics Check		OK	21:04
Air Blank		0.000	21:04
Control Test		0.078	21:05
Air Blank		0.000	21:05
Subject Sample	# 1	0.260	21:06
Breath Volume		1.910	21:06
Air Blank		0.000	21:06
Air Blank		0.000	21:08
Subject Sample	# 2	0.254	21:09
Breath Volume		1.812	21:09
Air Blank		0.000	21:09
Control Test		0.081	21:10
Air Blank		0.000	21:10
Diagnostics Check		OK	21:10

Data Download Date/Time: 12/21/2019 19:27

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MILLER,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: NOWELS

DATE/TIME OF TEST: 12/06/2019 22:03:19

Diagnostics Check		OK	22:00	
Air Blank		0.000	22:00	
Control Test		0.082	22:01	
Air Blank		0.000	22:01	
Subject Sample	# 1		22:01	Subject Test Refused
Breath Volume		0.000	22:01	
Air Blank		0.000	22:02	
Control Test		0.081	22:02	
Air Blank		0.000	22:03	
Diagnostics Check		OK	22:03	

Data Download Date/Time: 12/21/2019 19:27

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANWORP,JUDSON,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GROFF,RANDI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: MARTINEZ CALVA

DATE/TIME OF TEST: 12/07/2019 00:33:23

Diagnostics Check		OK	00:21	
Air Blank		0.000	00:22	
Control Test		0.082	00:22	
Air Blank		0.000	00:22	
Subject Sample	# 1	0.158	00:26	Volume Not Met
Breath Volume		0.144	00:26	
Air Blank		0.000	00:26	
Air Blank		0.000	00:28	
Subject Sample	# 2	0.160	00:31	Volume Not Met
Breath Volume		0.296	00:31	
Air Blank		0.000	00:32	
Control Test		0.079	00:32	
Air Blank		0.000	00:33	
Diagnostics Check		OK	00:33	

Data Download Date/Time: 12/21/2019 19:27

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANWORP,JUDSON,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GROFF,RANDI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: MARTINEZ CALVA

DATE/TIME OF TEST: 12/07/2019 00:43:49

Diagnostics Check		OK	00:35	
Air Blank		0.000	00:35	
Control Test		0.079	00:36	
Air Blank		0.000	00:36	
Subject Sample	# 1	0.156	00:39	Volume Not Met
Breath Volume		0.050	00:39	
Air Blank		0.000	00:40	
Air Blank		0.000	00:42	
Subject Sample	# 2		00:42	Subject Test Refused
Breath Volume		0.000	00:42	
Air Blank		0.000	00:42	
Control Test		0.080	00:43	
Air Blank		0.000	00:43	
Diagnostics Check		OK	00:43	

Data Download Date/Time: 12/21/2019 19:27

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: KIRK

DATE/TIME OF TEST: 12/07/2019 04:44:59

Diagnostics Check		OK	04:34	
Air Blank		0.000	04:35	
Control Test		0.082	04:35	
Air Blank		0.000	04:35	
Subject Sample	# 1		04:39	No Sample Provided
Breath Volume		0.000	04:39	
Air Blank		0.000	04:39	
Air Blank		0.000	04:41	
Subject Sample	# 2		04:43	Subject Test Refused
Breath Volume		0.000	04:43	
Air Blank		0.000	04:44	
Control Test		0.082	04:44	
Air Blank		0.000	04:44	
Diagnostics Check		OK	04:44	

Data Download Date/Time: 12/21/2019 19:27

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 05:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:26

SUBJECT'S LAST NAME: GONZALEZ-CANCEL

DATE/TIME OF TEST: 12/07/2019 05:53:11

Diagnostics Check		OK	05:47
Air Blank		0.000	05:47
Control Test		0.080	05:47
Air Blank		0.000	05:48
Subject Sample	# 1	0.224	05:48
Breath Volume		1.437	05:48
Air Blank		0.000	05:49
Air Blank		0.000	05:51
Subject Sample	# 2	0.231	05:51
Breath Volume		2.125	05:51
Air Blank		0.000	05:52
Control Test		0.080	05:52
Air Blank		0.000	05:53
Diagnostics Check		OK	05:53

Data Download Date/Time: 12/21/2019 19:27

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CLAGUE,E
ARRESTING OFFICER AGENCY: Gulfport PD
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: HENSON

DATE/TIME OF TEST: 12/08/2019 01:49:35

Diagnostics Check		OK	01:46	
Air Blank		0.000	01:47	
Control Test		0.082	01:47	
Air Blank		0.000	01:48	
Subject Sample	# 1		01:48	Subject Test Refused
Breath Volume		0.000	01:48	
Air Blank		0.000	01:48	
Control Test		0.081	01:49	
Air Blank		0.000	01:49	
Diagnostics Check		OK	01:49	

Data Download Date/Time: 12/21/2019 19:27

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 04:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: PONCE

DATE/TIME OF TEST: 12/08/2019 04:49:37

Diagnostics Check		OK	04:43
Air Blank		0.000	04:43
Control Test		0.082	04:44
Air Blank		0.000	04:44
Subject Sample	# 1	0.151	04:45
Breath Volume		2.468	04:45
Air Blank		0.000	04:45
Air Blank		0.000	04:47
Subject Sample	# 2	0.149	04:48
Breath Volume		2.964	04:48
Air Blank		0.000	04:48
Control Test		0.081	04:49
Air Blank		0.000	04:49
Diagnostics Check		OK	04:49

Data Download Date/Time: 12/21/2019 19:27

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELSHANS,DEAN,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HECK,WILLIAM
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: HAWKINS

DATE/TIME OF TEST: 12/09/2019 00:29:01

Diagnostics Check		OK	00:22
Air Blank		0.000	00:22
Control Test		0.081	00:23
Air Blank		0.000	00:23
Subject Sample	# 1	0.226	00:24
Breath Volume		1.921	00:24
Air Blank		0.000	00:25
Air Blank		0.000	00:26
Subject Sample	# 2	0.228	00:27
Breath Volume		1.601	00:27
Air Blank		0.000	00:28
Control Test		0.081	00:28
Air Blank		0.000	00:28
Diagnostics Check		OK	00:29

Data Download Date/Time: 12/21/2019 19:27

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SURITA,CHRISTIAN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SURITA,CHRISTIAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 19:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:06

SUBJECT'S LAST NAME: FRANTZIS

DATE/TIME OF TEST: 12/09/2019 20:52:42

Diagnostics Check		OK	20:46
Air Blank		0.000	20:47
Control Test		0.082	20:47
Air Blank		0.000	20:47
Subject Sample	# 1	0.229	20:48
Breath Volume		3.460	20:48
Air Blank		0.000	20:48
Air Blank		0.000	20:50
Subject Sample	# 2	0.243	20:51
Breath Volume		3.433	20:51
Air Blank		0.000	20:51
Control Test		0.080	20:52
Air Blank		0.000	20:52
Diagnostics Check		OK	20:52

Data Download Date/Time: 12/21/2019 19:28

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERSAUD,COMPTON,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PERSAUD,COMPTON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 21:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: YUHAS

DATE/TIME OF TEST: 12/09/2019 22:27:25

Diagnostics Check		OK	22:21
Air Blank		0.000	22:21
Control Test		0.081	22:22
Air Blank		0.000	22:22
Subject Sample	# 1	0.205	22:23
Breath Volume		1.769	22:23
Air Blank		0.000	22:23
Air Blank		0.000	22:25
Subject Sample	# 2	0.207	22:25
Breath Volume		1.375	22:25
Air Blank		0.000	22:26
Control Test		0.081	22:26
Air Blank		0.000	22:27
Diagnostics Check		OK	22:27

Data Download Date/Time: 12/21/2019 19:28

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERSAUD,COMPTON,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PERSAUD,COMPTON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: SENEY

DATE/TIME OF TEST: 12/10/2019 02:34:33

Diagnostics Check		OK	02:24
Air Blank		0.000	02:24
Control Test		0.082	02:25
Air Blank		0.000	02:25
Subject Sample	# 1	0.167	02:28
Breath Volume		1.492	02:28
Air Blank		0.000	02:29
Air Blank		0.000	02:30
Subject Sample	# 2	0.165	02:32
Breath Volume		1.664	02:32
Air Blank		0.000	02:33
Control Test		0.080	02:33
Air Blank		0.000	02:34
Diagnostics Check		OK	02:34

Data Download Date/Time: 12/21/2019 19:28

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERSAUD,COMPTON,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PERSAUD,COMPTON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: LE

DATE/TIME OF TEST: 12/11/2019 04:20:10

Diagnostics Check		OK	04:15	
Air Blank		0.000	04:15	
Control Test		0.082	04:16	
Air Blank		0.000	04:16	
Subject Sample	# 1		04:18	Subject Test Refused
Breath Volume		0.000	04:18	
Air Blank		0.000	04:19	
Control Test		0.081	04:19	
Air Blank		0.000	04:20	
Diagnostics Check		OK	04:20	

Data Download Date/Time: 12/21/2019 19:28

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICOTTILLI,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI,JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: DUROCHER

DATE/TIME OF TEST: 12/13/2019 01:18:44

Diagnostics Check		OK	01:12
Air Blank		0.000	01:13
Control Test		0.082	01:13
Air Blank		0.000	01:14
Subject Sample	# 1	0.000	01:14
Breath Volume		1.945	01:14
Air Blank		0.000	01:14
Air Blank		0.000	01:16
Subject Sample	# 2	0.000	01:17
Breath Volume		2.085	01:17
Air Blank		0.000	01:17
Control Test		0.081	01:18
Air Blank		0.000	01:18
Diagnostics Check		OK	01:18

Data Download Date/Time: 12/21/2019 19:28

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 19:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:06

SUBJECT'S LAST NAME: PAGUNI

DATE/TIME OF TEST: 12/13/2019 20:31:10

Diagnostics Check		OK	20:27	
Air Blank		0.000	20:28	
Control Test		0.082	20:28	
Air Blank		0.000	20:29	
Subject Sample	# 1		20:29	Subject Test Refused
Breath Volume		0.000	20:29	
Air Blank		0.000	20:30	
Control Test		0.081	20:30	
Air Blank		0.000	20:31	
Diagnostics Check		OK	20:31	

Data Download Date/Time: 12/21/2019 19:28

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERGER,CHARLES,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BERGER,CHARLES
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 04:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:16

SUBJECT'S LAST NAME: MARSHALL

DATE/TIME OF TEST: 12/14/2019 05:42:59

Diagnostics Check		OK	05:36
Air Blank		0.000	05:36
Control Test		0.082	05:37
Air Blank		0.000	05:37
Subject Sample	# 1	0.089	05:38
Breath Volume		2.003	05:38
Air Blank		0.000	05:38
Air Blank		0.000	05:40
Subject Sample	# 2	0.087	05:41
Breath Volume		2.132	05:41
Air Blank		0.000	05:42
Control Test		0.080	05:42
Air Blank		0.000	05:42
Diagnostics Check		OK	05:42

Data Download Date/Time: 12/21/2019 19:28

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 19:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:59

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 12/14/2019 20:24:14

Diagnostics Check		OK	20:21	
Air Blank		0.000	20:21	
Control Test		0.083	20:22	
Air Blank		0.000	20:22	
Subject Sample	# 1		20:22	Subject Test Refused
Breath Volume		0.000	20:22	
Air Blank		0.000	20:23	
Control Test		0.081	20:23	
Air Blank		0.000	20:24	
Diagnostics Check		OK	20:24	

Data Download Date/Time: 12/21/2019 19:28

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MULLINS,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 12/15/2019 00:03:02

Diagnostics Check		OK	00:00	
Air Blank		0.000	00:00	
Control Test		0.082	00:00	
Air Blank		0.000	00:01	
Subject Sample	# 1		00:01	Subject Test Refused
Breath Volume		0.000	00:01	
Air Blank		0.000	00:02	
Control Test		0.080	00:02	
Air Blank		0.000	00:02	
Diagnostics Check		OK	00:03	

Data Download Date/Time: 12/21/2019 19:28

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SCHULTZ,T
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: YOAKUM

DATE/TIME OF TEST: 12/15/2019 00:16:58

Diagnostics Check		OK	00:14	
Air Blank		0.000	00:14	
Control Test		0.081	00:14	
Air Blank		0.000	00:15	
Subject Sample	# 1		00:15	Subject Test Refused
Breath Volume		0.000	00:15	
Air Blank		0.000	00:16	
Control Test		0.081	00:16	
Air Blank		0.000	00:16	
Diagnostics Check		OK	00:16	

Data Download Date/Time: 12/21/2019 19:28

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BROWN,E
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: SPREITZER

DATE/TIME OF TEST: 12/15/2019 02:37:29

Diagnostics Check		OK	02:34	
Air Blank		0.000	02:35	
Control Test		0.082	02:35	
Air Blank		0.000	02:35	
Subject Sample	# 1		02:36	Subject Test Refused
Breath Volume		0.000	02:36	
Air Blank		0.000	02:36	
Control Test		0.081	02:36	
Air Blank		0.000	02:37	
Diagnostics Check		OK	02:37	

Data Download Date/Time: 12/21/2019 19:28

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: MELLOR

DATE/TIME OF TEST: 12/15/2019 04:12:02

Diagnostics Check		OK	04:09	
Air Blank		0.000	04:09	
Control Test		0.082	04:10	
Air Blank		0.000	04:10	
Subject Sample	# 1		04:10	Subject Test Refused
Breath Volume		0.000	04:10	
Air Blank		0.000	04:11	
Control Test		0.081	04:11	
Air Blank		0.000	04:11	
Diagnostics Check		OK	04:12	

Data Download Date/Time: 12/21/2019 19:28

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:56

SUBJECT'S LAST NAME: GOMBORONE

DATE/TIME OF TEST: 12/15/2019 21:23:05

Diagnostics Check		OK	21:20	
Air Blank		0.000	21:20	
Control Test		0.082	21:20	
Air Blank		0.000	21:21	
Subject Sample	# 1		21:21	Subject Test Refused
Breath Volume		0.000	21:21	
Air Blank		0.000	21:22	
Control Test		0.082	21:22	
Air Blank		0.000	21:22	
Diagnostics Check		OK	21:23	

Data Download Date/Time: 12/21/2019 19:28

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,NELSON,X
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MULLINS,JOHN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 20:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 12/15/2019 21:55:19

Diagnostics Check		OK	21:46	
Air Blank		0.000	21:47	
Control Test		0.082	21:47	
Air Blank		0.000	21:48	
Subject Sample	# 1		21:51	No Sample Provided
Breath Volume		0.000	21:51	
Air Blank		0.000	21:51	
Air Blank		0.000	21:53	
Subject Sample	# 2		21:53	Subject Test Refused
Breath Volume		0.000	21:53	
Air Blank		0.000	21:54	
Control Test		0.082	21:54	
Air Blank		0.000	21:55	
Diagnostics Check		OK	21:55	

Data Download Date/Time: 12/21/2019 19:28

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,WILLIAM,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: FAULKNER

DATE/TIME OF TEST: 12/16/2019 01:49:48

Diagnostics Check		OK	01:46	
Air Blank		0.000	01:47	
Control Test		0.082	01:47	
Air Blank		0.000	01:48	
Subject Sample	# 1		01:48	Subject Test Refused
Breath Volume		0.000	01:48	
Air Blank		0.000	01:48	
Control Test		0.081	01:49	
Air Blank		0.000	01:49	
Diagnostics Check		OK	01:49	

Data Download Date/Time: 12/21/2019 19:28

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:46

SUBJECT'S LAST NAME: CARON

DATE/TIME OF TEST: 12/16/2019 22:16:36

Diagnostics Check		OK	22:06	
Air Blank		0.000	22:07	
Control Test		0.082	22:07	
Air Blank		0.000	22:07	
Subject Sample	# 1	0.235	22:11	Volume Not Met
Breath Volume		0.062	22:11	
Air Blank		0.000	22:11	
Air Blank		0.000	22:13	
Subject Sample	# 2		22:14	Subject Test Refused
Breath Volume		0.062	22:14	
Air Blank		0.000	22:15	
Control Test		0.078	22:15	
Air Blank		0.000	22:16	
Diagnostics Check		OK	22:16	

Data Download Date/Time: 12/21/2019 19:28

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: NGUYEN,THUAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: TRACZEWSKI

DATE/TIME OF TEST: 12/16/2019 23:20:38

Diagnostics Check		OK	23:13
Air Blank		0.000	23:14
Control Test		0.081	23:14
Air Blank		0.000	23:15
Subject Sample	# 1	0.047	23:16
Breath Volume		1.640	23:16
Air Blank		0.000	23:16
Air Blank		0.000	23:18
Subject Sample	# 2	0.041	23:19
Breath Volume		1.203	23:19
Air Blank		0.000	23:19
Control Test		0.081	23:20
Air Blank		0.000	23:20
Diagnostics Check		OK	23:20

Data Download Date/Time: 12/21/2019 19:28

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSEN,STEPHEN,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HUFF,ZACHARY
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: SHERMAN

DATE/TIME OF TEST: 12/17/2019 01:37:03

Diagnostics Check		OK	01:31
Air Blank		0.000	01:31
Control Test		0.081	01:31
Air Blank		0.000	01:32
Subject Sample	# 1	0.000	01:32
Breath Volume		1.429	01:32
Air Blank		0.000	01:33
Air Blank		0.000	01:35
Subject Sample	# 2	0.000	01:35
Breath Volume		1.562	01:35
Air Blank		0.000	01:36
Control Test		0.081	01:36
Air Blank		0.000	01:36
Diagnostics Check		OK	01:37

Data Download Date/Time: 12/21/2019 19:28

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSEN,STEPHEN,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BOYLE,THOMAS
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: LASANE

DATE/TIME OF TEST: 12/17/2019 03:43:21

Diagnostics Check		OK	03:34	
Air Blank		0.000	03:35	
Control Test		0.082	03:35	
Air Blank		0.000	03:36	
Subject Sample	# 1	0.083	03:39	Volume Not Met
Breath Volume		0.285	03:39	
Air Blank		0.000	03:39	
Air Blank		0.000	03:41	
Subject Sample	# 2		03:41	Subject Test Refused
Breath Volume		0.000	03:41	
Air Blank		0.000	03:42	
Control Test		0.081	03:42	
Air Blank		0.000	03:43	
Diagnostics Check		OK	03:43	

Data Download Date/Time: 12/21/2019 19:28

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HUFF,Z
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 18:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:12

SUBJECT'S LAST NAME: SWANSON

DATE/TIME OF TEST: 12/17/2019 18:36:06

Diagnostics Check		OK	18:32	
Air Blank		0.000	18:33	
Control Test		0.082	18:33	
Air Blank		0.000	18:33	
Subject Sample	# 1		18:34	Subject Test Refused
Breath Volume		0.000	18:34	
Air Blank		0.000	18:35	
Control Test		0.081	18:35	
Air Blank		0.000	18:36	
Diagnostics Check		OK	18:36	

Data Download Date/Time: 12/21/2019 19:28

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ROLLESTON,J
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 19:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:56

SUBJECT'S LAST NAME: BRUCE

DATE/TIME OF TEST: 12/18/2019 20:21:28

Diagnostics Check		OK	20:18	
Air Blank		0.000	20:18	
Control Test		0.083	20:18	
Air Blank		0.000	20:19	
Subject Sample	# 1		20:20	Subject Test Refused
Breath Volume		0.000	20:20	
Air Blank		0.000	20:20	
Control Test		0.082	20:20	
Air Blank		0.000	20:21	
Diagnostics Check		OK	20:21	

Data Download Date/Time: 12/21/2019 19:28

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 12/18/2019 23:03:55

Diagnostics Check		OK	22:52	
Air Blank		0.000	22:52	
Control Test		0.082	22:53	
Air Blank		0.000	22:53	
Subject Sample	# 1		22:56	No Sample Provided
Breath Volume		0.000	22:56	
Air Blank		0.000	22:57	
Air Blank		0.000	22:59	
Subject Sample	# 2		23:02	No Sample Provided
Breath Volume		0.000	23:02	
Air Blank		0.000	23:03	
Control Test		0.083	23:03	
Air Blank		0.000	23:03	
Diagnostics Check		OK	23:03	

Data Download Date/Time: 12/21/2019 19:28

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 12/18/2019 23:19:12

Diagnostics Check		OK	23:16	
Air Blank		0.000	23:16	
Control Test		0.082	23:17	
Air Blank		0.000	23:17	
Subject Sample	# 1		23:17	Subject Test Refused
Breath Volume		0.000	23:17	
Air Blank		0.000	23:18	
Control Test		0.082	23:18	
Air Blank		0.000	23:19	
Diagnostics Check		OK	23:19	

Data Download Date/Time: 12/21/2019 19:28

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,WILLIAM,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 04:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:16

SUBJECT'S LAST NAME: OMALLEY-SIGLER

DATE/TIME OF TEST: 12/19/2019 05:43:39

Diagnostics Check		OK	05:37
Air Blank		0.000	05:37
Control Test		0.083	05:37
Air Blank		0.000	05:38
Subject Sample	# 1	0.277	05:39
Breath Volume		2.066	05:39
Air Blank		0.000	05:39
Air Blank		0.000	05:41
Subject Sample	# 2	0.259	05:42
Breath Volume		2.105	05:42
Air Blank		0.000	05:42
Control Test		0.082	05:43
Air Blank		0.000	05:43
Diagnostics Check		OK	05:43

Data Download Date/Time: 12/21/2019 19:28

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ROLLESTON,J
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: MITROVIC

DATE/TIME OF TEST: 12/19/2019 22:57:09

Diagnostics Check		OK	22:50
Air Blank		0.000	22:50
Control Test		0.083	22:51
Air Blank		0.000	22:51
Subject Sample	# 1	0.159	22:52
Breath Volume		2.035	22:52
Air Blank		0.000	22:52
Air Blank		0.000	22:54
Subject Sample	# 2	0.161	22:55
Breath Volume		2.000	22:55
Air Blank		0.000	22:56
Control Test		0.082	22:56
Air Blank		0.000	22:57
Diagnostics Check		OK	22:57

Data Download Date/Time: 12/21/2019 19:28

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ROLLESTON,J
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: KALTEUX

DATE/TIME OF TEST: 12/20/2019 01:47:13

Diagnostics Check		OK	01:39
Air Blank		0.000	01:39
Control Test		0.083	01:40
Air Blank		0.000	01:40
Subject Sample	# 1	0.103	01:42
Breath Volume		2.105	01:42
Air Blank		0.000	01:43
Air Blank		0.000	01:45
Subject Sample	# 2	0.106	01:45
Breath Volume		1.277	01:45
Air Blank		0.000	01:46
Control Test		0.081	01:46
Air Blank		0.000	01:47
Diagnostics Check		OK	01:47

Data Download Date/Time: 12/21/2019 19:28

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 04:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: RHEM

DATE/TIME OF TEST: 12/20/2019 05:29:24

Diagnostics Check		OK	05:22
Air Blank		0.000	05:22
Control Test		0.083	05:23
Air Blank		0.000	05:23
Subject Sample	# 1	0.129	05:24
Breath Volume		1.238	05:24
Air Blank		0.000	05:25
Air Blank		0.000	05:27
Subject Sample	# 2	0.132	05:27
Breath Volume		1.246	05:27
Air Blank		0.000	05:28
Control Test		0.081	05:28
Air Blank		0.000	05:29
Diagnostics Check		OK	05:29

Data Download Date/Time: 12/21/2019 19:28

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUDTKE,CHRISTOPHER,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LUDTKE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 19:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: FUSCO

DATE/TIME OF TEST: 12/20/2019 20:52:09

Diagnostics Check		OK	20:46
Air Blank		0.000	20:46
Control Test		0.082	20:46
Air Blank		0.000	20:47
Subject Sample	# 1	0.076	20:47
Breath Volume		1.789	20:47
Air Blank		0.000	20:48
Air Blank		0.000	20:50
Subject Sample	# 2	0.076	20:50
Breath Volume		2.265	20:50
Air Blank		0.000	20:51
Control Test		0.082	20:51
Air Blank		0.000	20:52
Diagnostics Check		OK	20:52

Data Download Date/Time: 12/21/2019 19:28

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: STUKUS,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:57

SUBJECT'S LAST NAME: CHAMBERLAIN

DATE/TIME OF TEST: 12/20/2019 21:29:13

Diagnostics Check		OK	21:22
Air Blank		0.000	21:22
Control Test		0.082	21:22
Air Blank		0.000	21:23
Subject Sample	# 1	0.177	21:24
Breath Volume		2.652	21:24
Air Blank		0.000	21:25
Air Blank		0.000	21:26
Subject Sample	# 2	0.172	21:27
Breath Volume		2.074	21:27
Air Blank		0.000	21:28
Control Test		0.080	21:28
Air Blank		0.000	21:29
Diagnostics Check		OK	21:29

Data Download Date/Time: 12/21/2019 19:28

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZSIDO,D
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 12/20/2019 22:31:20

Diagnostics Check		OK	22:18	
Air Blank		0.000	22:18	
Control Test		0.081	22:18	
Air Blank		0.000	22:19	
Subject Sample	# 1	0.164	22:22	Volume Not Met
Breath Volume		0.996	22:22	
Air Blank		0.000	22:23	
Air Blank		0.000	22:25	
Subject Sample	# 2	0.175	22:26	
Breath Volume		1.394	22:26	
Air Blank		0.000	22:27	
Air Blank		0.000	22:29	
Subject Sample	# 3	0.177	22:29	
Breath Volume		1.308	22:29	
Air Blank		0.000	22:30	
Control Test		0.082	22:30	
Air Blank		0.000	22:31	
Diagnostics Check		OK	22:31	

Data Download Date/Time: 12/21/2019 19:28

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: BRASH

DATE/TIME OF TEST: 12/21/2019 01:56:23

Diagnostics Check		OK	01:50
Air Blank		0.000	01:50
Control Test		0.082	01:50
Air Blank		0.000	01:51
Subject Sample	# 1	0.057	01:52
Breath Volume		2.046	01:52
Air Blank		0.000	01:52
Air Blank		0.000	01:54
Subject Sample	# 2	0.055	01:54
Breath Volume		1.578	01:54
Air Blank		0.000	01:55
Control Test		0.081	01:55
Air Blank		0.000	01:56
Diagnostics Check		OK	01:56

Data Download Date/Time: 12/21/2019 19:28

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SULLIVAN,T
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: LEVINE

DATE/TIME OF TEST: 12/21/2019 03:42:00

Diagnostics Check		OK	03:35
Air Blank		0.000	03:35
Control Test		0.082	03:36
Air Blank		0.000	03:36
Subject Sample	# 1	0.179	03:37
Breath Volume		1.933	03:37
Air Blank		0.000	03:38
Air Blank		0.000	03:39
Subject Sample	# 2	0.186	03:40
Breath Volume		1.851	03:40
Air Blank		0.000	03:41
Control Test		0.081	03:41
Air Blank		0.000	03:41
Diagnostics Check		OK	03:42

Data Download Date/Time: 12/21/2019 19:28

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEDE,BRENNEN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WEDE,BRENNEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: HARRINGTON

DATE/TIME OF TEST: 12/21/2019 04:30:26

Diagnostics Check		OK	04:22
Air Blank		0.000	04:23
Control Test		0.081	04:23
Air Blank		0.000	04:24
Subject Sample	# 1	0.125	04:24
Breath Volume		1.726	04:24
Air Blank		0.000	04:25
Air Blank		0.000	04:27
Subject Sample	# 2	0.125	04:28
Breath Volume		1.320	04:28
Air Blank		0.000	04:29
Control Test		0.082	04:29
Air Blank		0.000	04:30
Diagnostics Check		OK	04:30

Data Download Date/Time: 12/21/2019 19:28

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 04:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: P320784773410

DATE/TIME OF TEST: 12/21/2019 04:54:59

Diagnostics Check		OK	04:52	
Air Blank		0.000	04:52	
Control Test		0.081	04:52	
Air Blank		0.000	04:53	
Subject Sample	# 1		04:53	Subject Test Refused
Breath Volume		0.000	04:53	
Air Blank		0.000	04:54	
Control Test		0.081	04:54	
Air Blank		0.000	04:54	
Diagnostics Check		OK	04:54	

Data Download Date/Time: 12/21/2019 19:28

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORNING,PETER,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MACKAIL,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 07:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:19

SUBJECT'S LAST NAME: OTERO

DATE/TIME OF TEST: 12/21/2019 08:46:31

Diagnostics Check		OK	08:39
Air Blank		0.000	08:40
Control Test		0.083	08:40
Air Blank		0.000	08:41
Subject Sample	# 1	0.180	08:41
Breath Volume		1.718	08:41
Air Blank		0.000	08:42
Air Blank		0.000	08:44
Subject Sample	# 2	0.176	08:44
Breath Volume		1.214	08:44
Air Blank		0.000	08:45
Control Test		0.081	08:45
Air Blank		0.000	08:46
Diagnostics Check		OK	08:46

Data Download Date/Time: 12/21/2019 19:28

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORNING,PETER,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GRATT,JOSEPH
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 14:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:04

SUBJECT'S LAST NAME: LANEVILLE

DATE/TIME OF TEST: 12/21/2019 15:30:35

Diagnostics Check		OK	15:24
Air Blank		0.000	15:24
Control Test		0.082	15:25
Air Blank		0.000	15:25
Subject Sample	# 1	0.196	15:26
Breath Volume		1.980	15:26
Air Blank		0.000	15:26
Air Blank		0.000	15:28
Subject Sample	# 2	0.198	15:29
Breath Volume		1.976	15:29
Air Blank		0.000	15:29
Control Test		0.081	15:30
Air Blank		0.000	15:30
Diagnostics Check		OK	15:30

Data Download Date/Time: 12/21/2019 19:28

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ALVAREZ,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: GARCON

DATE/TIME OF TEST: 12/22/2019 00:09:59

Diagnostics Check		OK	00:07	
Air Blank		0.000	00:07	
Control Test		0.082	00:07	
Air Blank		0.000	00:08	
Subject Sample	# 1		00:08	Subject Test Refused
Breath Volume		0.000	00:08	
Air Blank		0.000	00:09	
Control Test		0.082	00:09	
Air Blank		0.000	00:09	
Diagnostics Check		OK	00:09	

Data Download Date/Time: 01/11/2020 20:33

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ROBB,S
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: MONNIG

DATE/TIME OF TEST: 12/22/2019 01:47:27

Diagnostics Check		OK	01:35	
Air Blank		0.000	01:36	
Control Test		0.082	01:36	
Air Blank		0.000	01:37	
Subject Sample	# 1		01:40	No Sample Provided
Breath Volume		0.000	01:40	
Air Blank		0.000	01:40	
Air Blank		0.000	01:42	
Subject Sample	# 2	0.172	01:45	Volume Not Met
Breath Volume		0.554	01:45	
Air Blank		0.000	01:46	
Control Test		0.081	01:46	
Air Blank		0.000	01:47	
Diagnostics Check		OK	01:47	

Data Download Date/Time: 01/11/2020 20:33

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZSIDO,D
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: MORLAND

DATE/TIME OF TEST: 12/22/2019 04:49:31

Diagnostics Check		OK	04:46	
Air Blank		0.000	04:47	
Control Test		0.081	04:47	
Air Blank		0.000	04:47	
Subject Sample	# 1		04:48	Subject Test Refused
Breath Volume		0.000	04:48	
Air Blank		0.000	04:48	
Control Test		0.081	04:48	
Air Blank		0.000	04:49	
Diagnostics Check		OK	04:49	

Data Download Date/Time: 01/11/2020 20:33

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSEN,STEPHEN,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: JACOBSEN,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 04:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: POWELL

DATE/TIME OF TEST: 12/22/2019 05:21:10

Diagnostics Check		OK	05:15
Air Blank		0.000	05:15
Control Test		0.080	05:15
Air Blank		0.000	05:16
Subject Sample	# 1	0.282	05:16
Breath Volume		2.621	05:16
Air Blank		0.000	05:17
Air Blank		0.000	05:19
Subject Sample	# 2	0.273	05:19
Breath Volume		2.371	05:19
Air Blank		0.000	05:20
Control Test		0.079	05:20
Air Blank		0.000	05:21
Diagnostics Check		OK	05:21

Data Download Date/Time: 01/11/2020 20:33

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,K,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CAREY,K
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 12/24/2019 03:12:19

Diagnostics Check		OK	03:05
Air Blank		0.000	03:05
Control Test		0.083	03:06
Air Blank		0.000	03:06
Subject Sample	# 1	0.000	03:07
Breath Volume		1.863	03:07
Air Blank		0.000	03:08
Air Blank		0.000	03:10
Subject Sample	# 2	0.000	03:10
Breath Volume		1.648	03:10
Air Blank		0.000	03:11
Control Test		0.082	03:11
Air Blank		0.000	03:12
Diagnostics Check		OK	03:12

Data Download Date/Time: 01/11/2020 20:33

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: HUNT

DATE/TIME OF TEST: 12/25/2019 02:39:59

Diagnostics Check		OK	02:32
Air Blank		0.000	02:32
Control Test		0.083	02:32
Air Blank		0.000	02:33
Subject Sample	# 1	0.149	02:34
Breath Volume		1.519	02:34
Air Blank		0.000	02:35
Air Blank		0.000	02:37
Subject Sample	# 2	0.153	02:38
Breath Volume		1.183	02:38
Air Blank		0.000	02:39
Control Test		0.080	02:39
Air Blank		0.000	02:39
Diagnostics Check		OK	02:39

Data Download Date/Time: 01/11/2020 20:33

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BYRD,W
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:56

SUBJECT'S LAST NAME: MCWILLIAMS

DATE/TIME OF TEST: 12/25/2019 20:30:06

Diagnostics Check		OK	20:18	
Air Blank		0.000	20:19	
Control Test		0.082	20:19	
Air Blank		0.000	20:20	
Subject Sample	# 1		20:23	No Sample Provided
Breath Volume		0.000	20:23	
Air Blank		0.000	20:23	
Air Blank		0.000	20:25	
Subject Sample	# 2		20:28	No Sample Provided
Breath Volume		0.000	20:28	
Air Blank		0.000	20:29	
Control Test		0.082	20:29	
Air Blank		0.000	20:30	
Diagnostics Check		OK	20:30	

Data Download Date/Time: 01/11/2020 20:33

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BYRD,W
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:22

SUBJECT'S LAST NAME: MCWILLIAMS

DATE/TIME OF TEST: 12/25/2019 20:51:16

Diagnostics Check		OK	20:43
Air Blank		0.000	20:43
Control Test		0.082	20:44
Air Blank		0.000	20:44
Subject Sample	# 1	0.157	20:46
Breath Volume		1.191	20:46
Air Blank		0.000	20:46
Air Blank		0.000	20:48
Subject Sample	# 2	0.153	20:49
Breath Volume		1.554	20:49
Air Blank		0.000	20:50
Control Test		0.081	20:50
Air Blank		0.000	20:51
Diagnostics Check		OK	20:51

Data Download Date/Time: 01/11/2020 20:33

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ROLLESTON,J
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:37

SUBJECT'S LAST NAME: LAPLANTE

DATE/TIME OF TEST: 12/26/2019 21:05:41

Diagnostics Check		OK	20:57
Air Blank		0.000	20:58
Control Test		0.082	20:58
Air Blank		0.000	20:58
Subject Sample	# 1	0.148	21:00
Breath Volume		2.300	21:00
Air Blank		0.000	21:01
Air Blank		0.000	21:03
Subject Sample	# 2	0.149	21:04
Breath Volume		2.488	21:04
Air Blank		0.000	21:04
Control Test		0.081	21:05
Air Blank		0.000	21:05
Diagnostics Check		OK	21:05

Data Download Date/Time: 01/11/2020 20:33

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERGER,CHARLES,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BERGER,CHARLES
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: SPENCER

DATE/TIME OF TEST: 12/27/2019 00:44:23

Diagnostics Check		OK	00:37
Air Blank		0.000	00:37
Control Test		0.082	00:38
Air Blank		0.000	00:38
Subject Sample	# 1	0.141	00:39
Breath Volume		2.589	00:39
Air Blank		0.000	00:40
Air Blank		0.000	00:42
Subject Sample	# 2	0.144	00:42
Breath Volume		2.347	00:42
Air Blank		0.000	00:43
Control Test		0.081	00:43
Air Blank		0.000	00:44
Diagnostics Check		OK	00:44

Data Download Date/Time: 01/11/2020 20:33

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: BATO

DATE/TIME OF TEST: 12/27/2019 03:15:05

Diagnostics Check		OK	03:08
Air Blank		0.000	03:09
Control Test		0.082	03:09
Air Blank		0.000	03:10
Subject Sample	# 1	0.139	03:10
Breath Volume		2.289	03:10
Air Blank		0.000	03:11
Air Blank		0.000	03:13
Subject Sample	# 2	0.142	03:13
Breath Volume		2.207	03:13
Air Blank		0.000	03:14
Control Test		0.081	03:14
Air Blank		0.000	03:14
Diagnostics Check		OK	03:15

Data Download Date/Time: 01/11/2020 20:33

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ROLLESTON,J
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: RODRIGUEZ-PEREZ

DATE/TIME OF TEST: 12/27/2019 23:56:46

Diagnostics Check		OK	23:50
Air Blank		0.000	23:50
Control Test		0.083	23:50
Air Blank		0.000	23:51
Subject Sample	# 1	0.147	23:52
Breath Volume		2.140	23:52
Air Blank		0.000	23:52
Air Blank		0.000	23:54
Subject Sample	# 2	0.141	23:55
Breath Volume		3.050	23:55
Air Blank		0.000	23:55
Control Test		0.080	23:56
Air Blank		0.000	23:56
Diagnostics Check		OK	23:56

Data Download Date/Time: 01/11/2020 20:33

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERGER,CHARLES,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BERGER,CHARLES
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: HORNER

DATE/TIME OF TEST: 12/28/2019 01:14:17

Diagnostics Check		OK	01:09	
Air Blank		0.000	01:10	
Control Test		0.082	01:10	
Air Blank		0.000	01:11	
Subject Sample	# 1		01:12	Subject Test Refused
Breath Volume		0.000	01:12	
Air Blank		0.000	01:13	
Control Test		0.081	01:13	
Air Blank		0.000	01:14	
Diagnostics Check		OK	01:14	

Data Download Date/Time: 01/11/2020 20:33

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: UPSON,IAN,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: UPSON

DATE/TIME OF TEST: 12/28/2019 03:07:15

Diagnostics Check		OK	03:00
Air Blank		0.000	03:01
Control Test		0.082	03:01
Air Blank		0.000	03:02
Subject Sample	# 1	0.171	03:02
Breath Volume		2.296	03:02
Air Blank		0.000	03:03
Air Blank		0.000	03:05
Subject Sample	# 2	0.177	03:05
Breath Volume		2.835	03:05
Air Blank		0.000	03:06
Control Test		0.081	03:06
Air Blank		0.000	03:07
Diagnostics Check		OK	03:07

Data Download Date/Time: 01/11/2020 20:33

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERGER,CHARLES,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BERGER,CHARLES
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 04:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: SPEARS

DATE/TIME OF TEST: 12/28/2019 05:37:28

Diagnostics Check		OK	05:30
Air Blank		0.000	05:30
Control Test		0.082	05:31
Air Blank		0.000	05:31
Subject Sample	# 1	0.170	05:32
Breath Volume		3.070	05:32
Air Blank		0.000	05:33
Air Blank		0.000	05:34
Subject Sample	# 2	0.162	05:35
Breath Volume		2.382	05:35
Air Blank		0.000	05:36
Control Test		0.081	05:36
Air Blank		0.000	05:37
Diagnostics Check		OK	05:37

Data Download Date/Time: 01/11/2020 20:33

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ROLLESTON,J
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: DUDIC

DATE/TIME OF TEST: 12/28/2019 21:02:30

Diagnostics Check		OK	20:55
Air Blank		0.000	20:56
Control Test		0.082	20:56
Air Blank		0.000	20:57
Subject Sample	# 1	0.219	20:57
Breath Volume		1.648	20:57
Air Blank		0.000	20:58
Air Blank		0.000	21:00
Subject Sample	# 2	0.216	21:00
Breath Volume		1.722	21:00
Air Blank		0.000	21:01
Control Test		0.079	21:01
Air Blank		0.000	21:02
Diagnostics Check		OK	21:02

Data Download Date/Time: 01/11/2020 20:33

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERGER,CHARLES,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BERGER,CHARLES
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: RAMIREZ DELGADO

DATE/TIME OF TEST: 12/29/2019 00:53:52

Diagnostics Check		OK	00:47
Air Blank		0.000	00:48
Control Test		0.082	00:48
Air Blank		0.000	00:48
Subject Sample	# 1	0.000	00:49
Breath Volume		4.148	00:49
Air Blank		0.000	00:49
Air Blank		0.000	00:51
Subject Sample	# 2	0.000	00:52
Breath Volume		3.437	00:52
Air Blank		0.000	00:52
Control Test		0.082	00:53
Air Blank		0.000	00:53
Diagnostics Check		OK	00:53

Data Download Date/Time: 01/11/2020 20:33

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DONOHUE,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: LEVI

DATE/TIME OF TEST: 12/29/2019 02:52:04

Diagnostics Check		OK	02:45
Air Blank		0.000	02:45
Control Test		0.082	02:45
Air Blank		0.000	02:46
Subject Sample	# 1	0.097	02:47
Breath Volume		2.515	02:47
Air Blank		0.000	02:47
Air Blank		0.000	02:49
Subject Sample	# 2	0.092	02:50
Breath Volume		1.953	02:50
Air Blank		0.000	02:51
Control Test		0.081	02:51
Air Blank		0.000	02:51
Diagnostics Check		OK	02:52

Data Download Date/Time: 01/11/2020 20:33

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: KNOLL-RUSSO

DATE/TIME OF TEST: 12/29/2019 04:22:29

Diagnostics Check		OK	04:15
Air Blank		0.000	04:16
Control Test		0.081	04:16
Air Blank		0.000	04:17
Subject Sample	# 1	0.118	04:17
Breath Volume		1.175	04:17
Air Blank		0.000	04:18
Air Blank		0.000	04:20
Subject Sample	# 2	0.123	04:20
Breath Volume		2.214	04:20
Air Blank		0.000	04:21
Control Test		0.080	04:21
Air Blank		0.000	04:22
Diagnostics Check		OK	04:22

Data Download Date/Time: 01/11/2020 20:33

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERSAUD,COMPTON,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PERSAUD,COMPTON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 18:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: RODGERS

DATE/TIME OF TEST: 12/29/2019 19:12:02

Diagnostics Check		OK	19:00	
Air Blank		0.000	19:01	
Control Test		0.083	19:01	
Air Blank		0.000	19:01	
Subject Sample	# 1		19:05	No Sample Provided
Breath Volume		0.000	19:05	
Air Blank		0.000	19:05	
Air Blank		0.000	19:07	
Subject Sample	# 2		19:10	No Sample Provided
Breath Volume		0.000	19:10	
Air Blank		0.000	19:11	
Control Test		0.082	19:11	
Air Blank		0.000	19:11	
Diagnostics Check		OK	19:12	

Data Download Date/Time: 01/11/2020 20:33

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EASTTY,MARK,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GUY,JOSH
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:57

SUBJECT'S LAST NAME: KALCK

DATE/TIME OF TEST: 12/29/2019 21:20:05

Diagnostics Check		OK	21:17	
Air Blank		0.000	21:17	
Control Test		0.082	21:18	
Air Blank		0.000	21:18	
Subject Sample	# 1		21:18	Subject Test Refused
Breath Volume		0.000	21:18	
Air Blank		0.000	21:19	
Control Test		0.081	21:19	
Air Blank		0.000	21:19	
Diagnostics Check		OK	21:20	

Data Download Date/Time: 01/11/2020 20:33

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERGER,CHARLES,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ROSADO,IAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: GILBERT

DATE/TIME OF TEST: 12/29/2019 22:30:11

Diagnostics Check		OK	22:21
Air Blank		0.000	22:22
Control Test		0.082	22:22
Air Blank		0.000	22:22
Subject Sample	# 1	0.242	22:25
Breath Volume		1.757	22:25
Air Blank		0.000	22:26
Air Blank		0.000	22:28
Subject Sample	# 2	0.251	22:28
Breath Volume		1.875	22:28
Air Blank		0.000	22:29
Control Test		0.079	22:29
Air Blank		0.000	22:30
Diagnostics Check		OK	22:30

Data Download Date/Time: 01/11/2020 20:33

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EASTTY,MARK,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MULLINS,JOHN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: MITSON

DATE/TIME OF TEST: 12/30/2019 00:12:47

Diagnostics Check		OK	00:03	
Air Blank		0.000	00:04	
Control Test		0.081	00:04	
Air Blank		0.000	00:05	
Subject Sample	# 1		00:08	No Sample Provided
Breath Volume		0.000	00:08	
Air Blank		0.000	00:08	
Air Blank		0.000	00:10	
Subject Sample	# 2		00:11	Subject Test Refused
Breath Volume		0.000	00:11	
Air Blank		0.000	00:11	
Control Test		0.081	00:12	
Air Blank		0.000	00:12	
Diagnostics Check		OK	00:12	

Data Download Date/Time: 01/11/2020 20:33

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EASTTY,MARK,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MULLINS,JOHN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: MITSON

DATE/TIME OF TEST: 12/30/2019 00:22:57

Diagnostics Check		OK	00:16
Air Blank		0.000	00:16
Control Test		0.082	00:17
Air Blank		0.000	00:17
Subject Sample	# 1	0.261	00:18
Breath Volume		1.496	00:18
Air Blank		0.000	00:19
Air Blank		0.000	00:21
Subject Sample	# 2	0.253	00:21
Breath Volume		1.679	00:21
Air Blank		0.000	00:22
Control Test		0.081	00:22
Air Blank		0.000	00:22
Diagnostics Check		OK	00:22

Data Download Date/Time: 01/11/2020 20:33

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EASTTY,MARK,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MACRINA,ANGELO
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: MUETZEL

DATE/TIME OF TEST: 12/30/2019 01:14:27

Diagnostics Check		OK	01:08
Air Blank		0.000	01:08
Control Test		0.081	01:08
Air Blank		0.000	01:09
Subject Sample	# 1	0.177	01:10
Breath Volume		2.273	01:10
Air Blank		0.000	01:10
Air Blank		0.000	01:12
Subject Sample	# 2	0.184	01:12
Breath Volume		2.195	01:12
Air Blank		0.000	01:13
Control Test		0.080	01:13
Air Blank		0.000	01:14
Diagnostics Check		OK	01:14

Data Download Date/Time: 01/11/2020 20:33

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERGER,CHARLES,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BERGER,CHARLES
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 04:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: HUTCHINSON

DATE/TIME OF TEST: 12/30/2019 05:31:53

Diagnostics Check		OK	05:25
Air Blank		0.000	05:25
Control Test		0.082	05:26
Air Blank		0.000	05:26
Subject Sample	# 1	0.154	05:27
Breath Volume		1.429	05:27
Air Blank		0.000	05:27
Air Blank		0.000	05:29
Subject Sample	# 2	0.161	05:30
Breath Volume		1.929	05:30
Air Blank		0.000	05:30
Control Test		0.081	05:31
Air Blank		0.000	05:31
Diagnostics Check		OK	05:31

Data Download Date/Time: 01/11/2020 20:33

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORNING,PETER,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HORNING,PETER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 17:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: LOGAN

DATE/TIME OF TEST: 12/30/2019 18:43:20

Diagnostics Check		OK	18:36
Air Blank		0.000	18:37
Control Test		0.082	18:37
Air Blank		0.000	18:37
Subject Sample	# 1	0.115	18:38
Breath Volume		1.941	18:38
Air Blank		0.000	18:39
Air Blank		0.000	18:40
Subject Sample	# 2	0.101	18:41
Breath Volume		1.488	18:41
Air Blank		0.000	18:42
Control Test		0.081	18:42
Air Blank		0.000	18:43
Diagnostics Check		OK	18:43

Data Download Date/Time: 01/11/2020 20:33

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PALERMITI,KATIE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 18:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:24

SUBJECT'S LAST NAME: BLACK

DATE/TIME OF TEST: 12/30/2019 19:50:33

Diagnostics Check		OK	19:47	
Air Blank		0.000	19:47	
Control Test		0.082	19:47	
Air Blank		0.000	19:48	
Subject Sample	# 1		19:49	Subject Test Refused
Breath Volume		0.000	19:49	
Air Blank		0.000	19:49	
Control Test		0.082	19:49	
Air Blank		0.000	19:50	
Diagnostics Check		OK	19:50	

Data Download Date/Time: 01/11/2020 20:33

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSEN,STEPHEN,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HUFF,ZACHARY
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 21:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: HARTMAN

DATE/TIME OF TEST: 12/30/2019 21:54:57

Diagnostics Check		OK	21:52	
Air Blank		0.000	21:52	
Control Test		0.081	21:52	
Air Blank		0.000	21:53	
Subject Sample	# 1		21:53	Subject Test Refused
Breath Volume		0.000	21:53	
Air Blank		0.000	21:54	
Control Test		0.081	21:54	
Air Blank		0.000	21:54	
Diagnostics Check		OK	21:54	

Data Download Date/Time: 01/11/2020 20:33

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DONOHUE,LAWRENCE
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: FLETCHER

DATE/TIME OF TEST: 12/31/2019 04:39:48

Diagnostics Check		OK	04:33
Air Blank		0.000	04:33
Control Test		0.082	04:34
Air Blank		0.000	04:34
Subject Sample	# 1	0.137	04:35
Breath Volume		1.777	04:35
Air Blank		0.000	04:35
Air Blank		0.000	04:37
Subject Sample	# 2	0.137	04:38
Breath Volume		1.726	04:38
Air Blank		0.000	04:38
Control Test		0.080	04:39
Air Blank		0.000	04:39
Diagnostics Check		OK	04:39

Data Download Date/Time: 01/11/2020 20:33

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AUSTIN,JOSEPH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: CURCI

DATE/TIME OF TEST: 12/31/2019 04:51:40

Diagnostics Check		OK	04:48	
Air Blank		0.000	04:49	
Control Test		0.081	04:49	
Air Blank		0.000	04:49	
Subject Sample	# 1		04:50	Subject Test Refused
Breath Volume		0.000	04:50	
Air Blank		0.000	04:50	
Control Test		0.081	04:51	
Air Blank		0.000	04:51	
Diagnostics Check		OK	04:51	

Data Download Date/Time: 01/11/2020 20:33

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: OREILLY,TERRENCE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: QUINN III

DATE/TIME OF TEST: 12/31/2019 05:24:09

Diagnostics Check		OK	05:20	
Air Blank		0.000	05:21	
Control Test		0.081	05:21	
Air Blank		0.000	05:22	
Subject Sample	# 1		05:22	Subject Test Refused
Breath Volume		0.000	05:22	
Air Blank		0.000	05:23	
Control Test		0.081	05:23	
Air Blank		0.000	05:24	
Diagnostics Check		OK	05:24	

Data Download Date/Time: 01/11/2020 20:33

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 22:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: KENNEDY

DATE/TIME OF TEST: 12/31/2019 23:07:18

Diagnostics Check		OK	23:00
Air Blank		0.000	23:00
Control Test		0.083	23:01
Air Blank		0.000	23:01
Subject Sample	# 1	0.106	23:02
Breath Volume		1.863	23:02
Air Blank		0.000	23:03
Air Blank		0.000	23:05
Subject Sample	# 2	0.108	23:05
Breath Volume		1.800	23:05
Air Blank		0.000	23:06
Control Test		0.081	23:06
Air Blank		0.000	23:07
Diagnostics Check		OK	23:07

Data Download Date/Time: 01/11/2020 20:33

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERIS-RUIZ,ELIZABETH,P
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: CORUJO,EMMANUEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 11/01/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 07:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:24

SUBJECT'S LAST NAME: STELLA

DATE/TIME OF TEST: 12/01/2019 07:57:36

Diagnostics Check		OK	07:50
Air Blank		0.000	07:51
Control Test		0.079	07:51
Air Blank		0.000	07:52
Subject Sample	# 1	0.153	07:53
Breath Volume		2.406	07:53
Air Blank		0.000	07:53
Air Blank		0.000	07:55
Subject Sample	# 2	0.155	07:55
Breath Volume		3.636	07:55
Air Blank		0.000	07:56
Control Test		0.078	07:57
Air Blank		0.000	07:57
Diagnostics Check		OK	07:57

Data Download Date/Time: 12/26/2019 19:33

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HURT,JONATHAN,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: HURT,JONATHAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 11/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:54

SUBJECT'S LAST NAME: PESCE

DATE/TIME OF TEST: 12/04/2019 20:26:14

Diagnostics Check		OK	20:16
Air Blank		0.000	20:17
Control Test		0.080	20:17
Air Blank		0.000	20:17
Subject Sample	# 1	0.167	20:20
Breath Volume		1.730	20:20
Air Blank		0.000	20:21
Air Blank		0.000	20:22
Subject Sample	# 2	0.185	20:24
Breath Volume		1.558	20:24
Air Blank		0.000	20:25
Control Test		0.077	20:25
Air Blank		0.000	20:26
Diagnostics Check		OK	20:26

Data Download Date/Time: 12/26/2019 19:33

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VENEZIA,STEVEN,L
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: VENEZIA,STEVEN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 11/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: BALLENTINE

DATE/TIME OF TEST: 12/07/2019 00:13:45

Diagnostics Check		OK	00:07
Air Blank		0.000	00:07
Control Test		0.079	00:08
Air Blank		0.000	00:08
Subject Sample	# 1	0.189	00:09
Breath Volume		2.023	00:09
Air Blank		0.000	00:09
Air Blank		0.000	00:11
Subject Sample	# 2	0.191	00:12
Breath Volume		2.023	00:12
Air Blank		0.000	00:12
Control Test		0.080	00:13
Air Blank		0.000	00:13
Diagnostics Check		OK	00:13

Data Download Date/Time: 12/26/2019 19:33

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HURT,JONATHAN,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: HURT,JONATHAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 11/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: ABREU

DATE/TIME OF TEST: 12/07/2019 01:49:19

Diagnostics Check		OK	01:45	
Air Blank		0.000	01:45	
Control Test		0.079	01:46	
Air Blank		0.000	01:46	
Subject Sample	# 1		01:47	RFI Detect
Breath Volume		0.000	01:47	
Air Blank		0.000	01:48	
Control Test		0.079	01:48	
Air Blank		0.000	01:49	
Diagnostics Check		OK	01:49	

Data Download Date/Time: 12/26/2019 19:33

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HURT,JONATHAN,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: HURT,JONATHAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 11/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: ABREU

DATE/TIME OF TEST: 12/07/2019 01:59:21

Diagnostics Check		OK	01:52
Air Blank		0.000	01:52
Control Test		0.079	01:52
Air Blank		0.000	01:53
Subject Sample	# 1	0.063	01:54
Breath Volume		3.234	01:54
Air Blank		0.000	01:55
Air Blank		0.000	01:56
Subject Sample	# 2	0.063	01:57
Breath Volume		3.464	01:57
Air Blank		0.000	01:58
Control Test		0.080	01:58
Air Blank		0.000	01:59
Diagnostics Check		OK	01:59

Data Download Date/Time: 12/26/2019 19:33

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEGERSMITH,DANIEL,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,DANIEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 11/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: PINILLA

DATE/TIME OF TEST: 12/07/2019 04:37:56

Diagnostics Check		OK	04:30
Air Blank		0.000	04:31
Control Test		0.079	04:31
Air Blank		0.000	04:32
Subject Sample	# 1	0.181	04:33
Breath Volume		2.238	04:33
Air Blank		0.000	04:33
Air Blank		0.000	04:35
Subject Sample	# 2	0.174	04:36
Breath Volume		2.355	04:36
Air Blank		0.000	04:37
Control Test		0.078	04:37
Air Blank		0.000	04:37
Diagnostics Check		OK	04:37

Data Download Date/Time: 12/26/2019 19:33

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEGERSMITH,DANIEL,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,NEGERSMITH
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 11/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: MCCLORY

DATE/TIME OF TEST: 12/08/2019 03:28:05

Diagnostics Check		OK	03:16	
Air Blank		0.000	03:17	
Control Test		0.079	03:17	
Air Blank		0.000	03:17	
Subject Sample	# 1		03:21	No Sample Provided
Breath Volume		0.000	03:21	
Air Blank		0.000	03:21	
Air Blank		0.000	03:23	
Subject Sample	# 2		03:26	No Sample Provided
Breath Volume		0.000	03:26	
Air Blank		0.000	03:27	
Control Test		0.079	03:27	
Air Blank		0.000	03:28	
Diagnostics Check		OK	03:28	

Data Download Date/Time: 12/26/2019 19:33

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEGERSMITH,DANIEL,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,DANIEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 11/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:42

SUBJECT'S LAST NAME: URTIAGA

DATE/TIME OF TEST: 12/12/2019 20:10:18

Diagnostics Check		OK	20:02
Air Blank		0.000	20:03
Control Test		0.079	20:03
Air Blank		0.000	20:04
Subject Sample	# 1	0.081	20:05
Breath Volume		1.914	20:05
Air Blank		0.000	20:05
Air Blank		0.000	20:07
Subject Sample	# 2	0.075	20:08
Breath Volume		1.785	20:08
Air Blank		0.000	20:09
Control Test		0.079	20:09
Air Blank		0.000	20:10
Diagnostics Check		OK	20:10

Data Download Date/Time: 12/26/2019 19:33

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEGERSMITH,DANIEL,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,DANIEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 11/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 12/13/2019 03:01:15

Diagnostics Check		OK	02:47	
Air Blank		0.000	02:48	
Control Test		0.079	02:48	
Air Blank		0.000	02:48	
Subject Sample	# 1	0.117	02:52	Volume Not Met
Breath Volume		0.828	02:52	
Air Blank		0.000	02:52	
Air Blank		0.000	02:54	
Subject Sample	# 2	0.121	02:56	
Breath Volume		2.242	02:56	
Air Blank		0.000	02:56	
Air Blank		0.000	02:58	
Subject Sample	# 3	0.127	02:59	
Breath Volume		2.355	02:59	
Air Blank		0.000	03:00	
Control Test		0.078	03:00	
Air Blank		0.000	03:01	
Diagnostics Check		OK	03:01	

Data Download Date/Time: 12/26/2019 19:33

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HURT,JONATHAN,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: HURT,JONATHAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 11/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:48

SUBJECT'S LAST NAME: LOPEZ SALAZAR

DATE/TIME OF TEST: 12/13/2019 21:26:39

Diagnostics Check		OK	21:18
Air Blank		0.000	21:19
Control Test		0.079	21:19
Air Blank		0.000	21:20
Subject Sample	# 1	0.084	21:21
Breath Volume		1.613	21:21
Air Blank		0.000	21:22
Air Blank		0.000	21:24
Subject Sample	# 2	0.092	21:25
Breath Volume		2.332	21:25
Air Blank		0.000	21:25
Control Test		0.079	21:26
Air Blank		0.000	21:26
Diagnostics Check		OK	21:26

Data Download Date/Time: 12/26/2019 19:33

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HURT,JONATHAN,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: HURT,JONATHAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 11/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 12/14/2019 02:23:42

Diagnostics Check		OK	02:15
Air Blank		0.000	02:16
Control Test		0.079	02:16
Air Blank		0.000	02:16
Subject Sample	# 1	0.227	02:18
Breath Volume		2.960	02:18
Air Blank		0.000	02:19
Air Blank		0.000	02:21
Subject Sample	# 2	0.230	02:22
Breath Volume		3.035	02:22
Air Blank		0.000	02:22
Control Test		0.077	02:23
Air Blank		0.000	02:23
Diagnostics Check		OK	02:23

Data Download Date/Time: 12/26/2019 19:33

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STIERS,S,T
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: STIERS,S
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 11/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 05:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:07

SUBJECT'S LAST NAME: LAMBERTSON

DATE/TIME OF TEST: 12/14/2019 06:01:56

Diagnostics Check		OK	05:46	
Air Blank		0.000	05:46	
Control Test		0.079	05:46	
Air Blank		0.000	05:47	
Subject Sample	# 1	0.189	05:50	Volume Not Met
Breath Volume		0.355	05:50	
Air Blank		0.000	05:51	
Air Blank		0.000	05:52	
Subject Sample	# 2	0.230	05:54	
Breath Volume		1.898	05:54	
Air Blank		0.000	05:55	
Air Blank		0.000	05:57	
Subject Sample	# 3	0.185	06:00	Volume Not Met
Breath Volume		0.683	06:00	
Air Blank		0.000	06:00	
Control Test		0.077	06:01	
Air Blank		0.000	06:01	
Diagnostics Check		OK	06:01	

Data Download Date/Time: 12/26/2019 19:33

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEGERSMITH,DANIEL,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,DANIEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 11/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:16

SUBJECT'S LAST NAME: GRANT

DATE/TIME OF TEST: 12/14/2019 22:56:10

Diagnostics Check		OK	22:44	
Air Blank		0.000	22:45	
Control Test		0.079	22:45	
Air Blank		0.000	22:46	
Subject Sample	# 1		22:49	No Sample Provided
Breath Volume		0.000	22:49	
Air Blank		0.000	22:49	
Air Blank		0.000	22:51	
Subject Sample	# 2		22:54	No Sample Provided
Breath Volume		0.000	22:54	
Air Blank		0.000	22:55	
Control Test		0.079	22:55	
Air Blank		0.000	22:56	
Diagnostics Check		OK	22:56	

Data Download Date/Time: 12/26/2019 19:34

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERG,BRITTANI,J
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: DURRANCE,PAUL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 11/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 10:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:09

SUBJECT'S LAST NAME: HUBBLE

DATE/TIME OF TEST: 12/15/2019 11:33:10

Diagnostics Check		OK	11:30	
Air Blank		0.000	11:30	
Control Test		0.079	11:30	
Air Blank		0.000	11:31	
Subject Sample	# 1		11:31	Subject Test Refused
Breath Volume		0.000	11:31	
Air Blank		0.000	11:32	
Control Test		0.079	11:32	
Air Blank		0.000	11:33	
Diagnostics Check		OK	11:33	

Data Download Date/Time: 12/26/2019 19:34

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUIS,OFC. J,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: KURINSKY,OFC.D
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 11/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 12/16/2019 23:03:52

Diagnostics Check		OK	22:57
Air Blank		0.000	22:57
Control Test		0.079	22:58
Air Blank		0.000	22:58
Subject Sample	# 1	0.043	22:59
Breath Volume		2.457	22:59
Air Blank		0.000	22:59
Air Blank		0.000	23:01
Subject Sample	# 2	0.040	23:02
Breath Volume		1.988	23:02
Air Blank		0.000	23:02
Control Test		0.079	23:03
Air Blank		0.000	23:03
Diagnostics Check		OK	23:03

Data Download Date/Time: 12/26/2019 19:34

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,CHAD,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: REED,CHAD
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 11/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 12/18/2019 22:53:19

Diagnostics Check		OK	22:46
Air Blank		0.000	22:46
Control Test		0.079	22:47
Air Blank		0.000	22:47
Subject Sample	# 1	0.162	22:48
Breath Volume		1.347	22:48
Air Blank		0.000	22:48
Air Blank		0.000	22:50
Subject Sample	# 2	0.149	22:51
Breath Volume		1.679	22:51
Air Blank		0.000	22:52
Control Test		0.079	22:52
Air Blank		0.000	22:53
Diagnostics Check		OK	22:53

Data Download Date/Time: 12/26/2019 19:34

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUIS,OFC.J,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: BUIS,OFC. J
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 12/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: FEIST

DATE/TIME OF TEST: 12/21/2019 03:20:23

Diagnostics Check		OK	03:13
Air Blank		0.000	03:13
Control Test		0.079	03:14
Air Blank		0.000	03:14
Subject Sample	# 1	0.090	03:15
Breath Volume		3.171	03:15
Air Blank		0.000	03:16
Air Blank		0.000	03:18
Subject Sample	# 2	0.080	03:18
Breath Volume		2.402	03:18
Air Blank		0.000	03:19
Control Test		0.078	03:19
Air Blank		0.000	03:20
Diagnostics Check		OK	03:20

Data Download Date/Time: 12/26/2019 19:34

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEGERSMITH,DANIEL,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,DANIEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 12/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:29

SUBJECT'S LAST NAME: MENDOZA

DATE/TIME OF TEST: 12/21/2019 21:01:57

Diagnostics Check		OK	20:54
Air Blank		0.000	20:55
Control Test		0.079	20:55
Air Blank		0.000	20:56
Subject Sample	# 1	0.088	20:57
Breath Volume		2.910	20:57
Air Blank		0.000	20:57
Air Blank		0.000	20:59
Subject Sample	# 2	0.083	21:00
Breath Volume		2.660	21:00
Air Blank		0.000	21:01
Control Test		0.079	21:01
Air Blank		0.000	21:01
Diagnostics Check		OK	21:01

Data Download Date/Time: 12/26/2019 19:34

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEGERSMITH,DANIEL,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,DANIEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 12/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: DEVARS

DATE/TIME OF TEST: 12/22/2019 04:14:40

Diagnostics Check		OK	04:06
Air Blank		0.000	04:07
Control Test		0.079	04:07
Air Blank		0.000	04:08
Subject Sample	# 1	0.000	04:09
Breath Volume		1.253	04:09
Air Blank		0.000	04:09
Air Blank		0.000	04:11
Subject Sample	# 2	0.000	04:13
Breath Volume		1.210	04:13
Air Blank		0.000	04:13
Control Test		0.079	04:14
Air Blank		0.000	04:14
Diagnostics Check		OK	04:14

Data Download Date/Time: 12/26/2019 19:34

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARNES,BRYAN,J
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: KLEIN,KEVIN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 12/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 08:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:56

SUBJECT'S LAST NAME: TYSON

DATE/TIME OF TEST: 12/22/2019 09:26:58

Diagnostics Check		OK	09:17
Air Blank		0.000	09:17
Control Test		0.079	09:17
Air Blank		0.000	09:18
Subject Sample	# 1	0.081	09:21
Breath Volume		1.976	09:21
Air Blank		0.000	09:21
Air Blank		0.000	09:23
Subject Sample	# 2	0.084	09:25
Breath Volume		1.648	09:25
Air Blank		0.000	09:26
Control Test		0.079	09:26
Air Blank		0.000	09:26
Diagnostics Check		OK	09:26

Data Download Date/Time: 12/26/2019 19:34

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VENEZIA,STEVEN,L
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: VENEZIA,STEVEN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 12/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:34

SUBJECT'S LAST NAME: OLEARY

DATE/TIME OF TEST: 12/22/2019 21:00:52

Diagnostics Check		OK	20:54
Air Blank		0.000	20:55
Control Test		0.079	20:55
Air Blank		0.000	20:55
Subject Sample	# 1	0.153	20:56
Breath Volume		1.808	20:56
Air Blank		0.000	20:56
Air Blank		0.000	20:58
Subject Sample	# 2	0.158	20:59
Breath Volume		2.847	20:59
Air Blank		0.000	20:59
Control Test		0.078	21:00
Air Blank		0.000	21:00
Diagnostics Check		OK	21:00

Data Download Date/Time: 12/26/2019 19:34

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEGERSMITH,DANIEL,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,DANIEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 12/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: LUTHER

DATE/TIME OF TEST: 12/27/2019 01:23:19

Diagnostics Check		OK	01:16
Air Blank		0.000	01:17
Control Test		0.080	01:17
Air Blank		0.000	01:18
Subject Sample	# 1	0.000	01:18
Breath Volume		2.285	01:18
Air Blank		0.000	01:19
Air Blank		0.000	01:21
Subject Sample	# 2	0.000	01:21
Breath Volume		1.902	01:21
Air Blank		0.000	01:22
Control Test		0.080	01:22
Air Blank		0.000	01:23
Diagnostics Check		OK	01:23

Data Download Date/Time: 02/03/2020 18:13

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEGERSMITH,DANIEL,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: BANKS,GARRETT
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 12/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:04

SUBJECT'S LAST NAME: MENESES LUNA

DATE/TIME OF TEST: 12/27/2019 19:36:35

Diagnostics Check		OK	19:25	
Air Blank		0.000	19:25	
Control Test		0.079	19:25	
Air Blank		0.000	19:26	
Subject Sample	# 1		19:29	No Sample Provided
Breath Volume		0.000	19:29	
Air Blank		0.000	19:30	
Air Blank		0.000	19:32	
Subject Sample	# 2		19:35	No Sample Provided
Breath Volume		0.000	19:35	
Air Blank		0.000	19:35	
Control Test		0.079	19:35	
Air Blank		0.000	19:36	
Diagnostics Check		OK	19:36	

Data Download Date/Time: 02/03/2020 18:13

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LASCKO,J,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: LASCKO,J
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 12/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: SANDERS

DATE/TIME OF TEST: 12/29/2019 02:29:47

Diagnostics Check		OK	02:18	
Air Blank		0.000	02:18	
Control Test		0.079	02:19	
Air Blank		0.000	02:19	
Subject Sample	# 1	0.083	02:22	Volume Not Met
Breath Volume		0.261	02:22	
Air Blank		0.000	02:23	
Air Blank		0.000	02:25	
Subject Sample	# 2		02:28	No Sample Provided
Breath Volume		0.000	02:28	
Air Blank		0.000	02:28	
Control Test		0.078	02:29	
Air Blank		0.000	02:29	
Diagnostics Check		OK	02:29	

Data Download Date/Time: 02/03/2020 18:13

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: GIBSON

DATE/TIME OF TEST: 12/01/2019 03:32:07

Diagnostics Check		OK	03:24
Air Blank		0.000	03:25
Control Test		0.081	03:25
Air Blank		0.000	03:26
Subject Sample	# 1	0.100	03:27
Breath Volume		1.500	03:27
Air Blank		0.000	03:28
Air Blank		0.000	03:29
Subject Sample	# 2	0.102	03:30
Breath Volume		1.425	03:30
Air Blank		0.000	03:31
Control Test		0.081	03:31
Air Blank		0.000	03:32
Diagnostics Check		OK	03:32

Data Download Date/Time: 12/20/2019 10:28

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVERETT,I,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: REYNOLDS,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 07:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:52

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 12/01/2019 09:20:46

Diagnostics Check		OK	09:14
Air Blank		0.000	09:15
Control Test		0.081	09:15
Air Blank		0.000	09:15
Subject Sample	# 1	0.171	09:16
Breath Volume		2.433	09:16
Air Blank		0.000	09:16
Air Blank		0.000	09:18
Subject Sample	# 2	0.172	09:19
Breath Volume		2.171	09:19
Air Blank		0.000	09:19
Control Test		0.080	09:20
Air Blank		0.000	09:20
Diagnostics Check		OK	09:20

Data Download Date/Time: 12/20/2019 10:28

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: WRIGHY

DATE/TIME OF TEST: 12/02/2019 00:41:20

Diagnostics Check		OK	00:34
Air Blank		0.000	00:35
Control Test		0.081	00:35
Air Blank		0.000	00:36
Subject Sample	# 1	0.189	00:36
Breath Volume		4.046	00:36
Air Blank		0.000	00:37
Air Blank		0.000	00:39
Subject Sample	# 2	0.183	00:39
Breath Volume		4.582	00:39
Air Blank		0.000	00:40
Control Test		0.080	00:40
Air Blank		0.000	00:41
Diagnostics Check		OK	00:41

Data Download Date/Time: 12/20/2019 10:28

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 12/02/2019 04:16:32

Diagnostics Check		OK	04:13	
Air Blank		0.000	04:14	
Control Test		0.081	04:14	
Air Blank		0.000	04:14	
Subject Sample	# 1		04:15	Subject Test Refused
Breath Volume		0.000	04:15	
Air Blank		0.000	04:15	
Control Test		0.082	04:15	
Air Blank		0.000	04:16	
Diagnostics Check		OK	04:16	

Data Download Date/Time: 12/20/2019 10:28

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: MIDCAP

DATE/TIME OF TEST: 12/03/2019 03:17:08

Diagnostics Check		OK	03:10
Air Blank		0.000	03:11
Control Test		0.083	03:11
Air Blank		0.000	03:11
Subject Sample	# 1	0.124	03:12
Breath Volume		3.964	03:12
Air Blank		0.000	03:13
Air Blank		0.000	03:14
Subject Sample	# 2	0.131	03:15
Breath Volume		3.972	03:15
Air Blank		0.000	03:16
Control Test		0.082	03:16
Air Blank		0.000	03:17
Diagnostics Check		OK	03:17

Data Download Date/Time: 12/20/2019 10:28

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,D,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TISBY,K
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 08:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:40

SUBJECT'S LAST NAME: EASTER

DATE/TIME OF TEST: 12/03/2019 10:05:08

Diagnostics Check		OK	10:02	
Air Blank		0.000	10:02	
Control Test		0.083	10:02	
Air Blank		0.000	10:03	
Subject Sample	# 1		10:03	Subject Test Refused
Breath Volume		0.000	10:03	
Air Blank		0.000	10:04	
Control Test		0.083	10:04	
Air Blank		0.000	10:05	
Diagnostics Check		OK	10:05	

Data Download Date/Time: 12/20/2019 10:28

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,M,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: WAHL

DATE/TIME OF TEST: 12/04/2019 01:10:05

Diagnostics Check		OK	01:03
Air Blank		0.000	01:04
Control Test		0.083	01:04
Air Blank		0.000	01:05
Subject Sample	# 1	0.007	01:05
Breath Volume		3.292	01:05
Air Blank		0.000	01:06
Air Blank		0.000	01:08
Subject Sample	# 2	0.007	01:08
Breath Volume		2.839	01:08
Air Blank		0.000	01:09
Control Test		0.083	01:09
Air Blank		0.000	01:10
Diagnostics Check		OK	01:10

Data Download Date/Time: 12/20/2019 10:28

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,M,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: MATHIS

DATE/TIME OF TEST: 12/04/2019 23:58:22

Diagnostics Check		OK	23:55	
Air Blank		0.000	23:55	
Control Test		0.082	23:56	
Air Blank		0.000	23:56	
Subject Sample	# 1		23:56	Subject Test Refused
Breath Volume		0.000	23:56	
Air Blank		0.000	23:57	
Control Test		0.081	23:57	
Air Blank		0.000	23:58	
Diagnostics Check		OK	23:58	

Data Download Date/Time: 12/20/2019 10:28

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,M,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: WERLING

DATE/TIME OF TEST: 12/06/2019 02:14:04

Diagnostics Check		OK	02:07
Air Blank		0.000	02:07
Control Test		0.083	02:08
Air Blank		0.000	02:08
Subject Sample	# 1	0.060	02:09
Breath Volume		1.925	02:09
Air Blank		0.000	02:09
Air Blank		0.000	02:11
Subject Sample	# 2	0.059	02:12
Breath Volume		1.914	02:12
Air Blank		0.000	02:13
Control Test		0.082	02:13
Air Blank		0.000	02:13
Diagnostics Check		OK	02:14

Data Download Date/Time: 12/20/2019 10:28

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,M,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: HARRELL

DATE/TIME OF TEST: 12/06/2019 03:54:31

Diagnostics Check		OK	03:47
Air Blank		0.000	03:47
Control Test		0.082	03:47
Air Blank		0.000	03:48
Subject Sample	# 1	0.179	03:49
Breath Volume		1.984	03:49
Air Blank		0.000	03:49
Air Blank		0.000	03:51
Subject Sample	# 2	0.187	03:52
Breath Volume		1.546	03:52
Air Blank		0.000	03:53
Control Test		0.082	03:53
Air Blank		0.000	03:54
Diagnostics Check		OK	03:54

Data Download Date/Time: 12/20/2019 10:28

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,M,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMONS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 12/06/2019 22:46:59

Diagnostics Check		OK	22:43	
Air Blank		0.000	22:44	
Control Test		0.082	22:44	
Air Blank		0.000	22:45	
Subject Sample	# 1		22:45	Subject Test Refused
Breath Volume		0.000	22:45	
Air Blank		0.000	22:46	
Control Test		0.082	22:46	
Air Blank		0.000	22:46	
Diagnostics Check		OK	22:46	

Data Download Date/Time: 12/20/2019 10:28

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,M,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: ROBERSON

DATE/TIME OF TEST: 12/07/2019 01:42:55

Diagnostics Check		OK	01:36
Air Blank		0.000	01:36
Control Test		0.082	01:36
Air Blank		0.000	01:37
Subject Sample	# 1	0.059	01:38
Breath Volume		1.140	01:38
Air Blank		0.000	01:38
Air Blank		0.000	01:40
Subject Sample	# 2	0.058	01:41
Breath Volume		1.640	01:41
Air Blank		0.000	01:42
Control Test		0.081	01:42
Air Blank		0.000	01:42
Diagnostics Check		OK	01:42

Data Download Date/Time: 12/20/2019 10:28

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,M,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAWSON,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: WEEKS

DATE/TIME OF TEST: 12/07/2019 03:05:01

Diagnostics Check		OK	02:58
Air Blank		0.000	02:58
Control Test		0.082	02:59
Air Blank		0.000	02:59
Subject Sample	# 1	0.164	03:00
Breath Volume		2.023	03:00
Air Blank		0.000	03:01
Air Blank		0.000	03:02
Subject Sample	# 2	0.158	03:03
Breath Volume		1.750	03:03
Air Blank		0.000	03:04
Control Test		0.081	03:04
Air Blank		0.000	03:04
Diagnostics Check		OK	03:05

Data Download Date/Time: 12/20/2019 10:28

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,M,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FILBERT,E
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: KOPF

DATE/TIME OF TEST: 12/07/2019 03:49:55

Diagnostics Check		OK	03:42
Air Blank		0.000	03:42
Control Test		0.082	03:43
Air Blank		0.000	03:43
Subject Sample	# 1	0.115	03:45
Breath Volume		1.519	03:45
Air Blank		0.000	03:45
Air Blank		0.000	03:47
Subject Sample	# 2	0.121	03:48
Breath Volume		2.023	03:48
Air Blank		0.000	03:49
Control Test		0.080	03:49
Air Blank		0.000	03:49
Diagnostics Check		OK	03:49

Data Download Date/Time: 12/20/2019 10:28

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,M,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PITTMAN,K
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: COLEMAN

DATE/TIME OF TEST: 12/07/2019 04:36:42

Diagnostics Check		OK	04:30
Air Blank		0.000	04:30
Control Test		0.081	04:30
Air Blank		0.000	04:31
Subject Sample	# 1	0.172	04:31
Breath Volume		1.183	04:31
Air Blank		0.000	04:32
Air Blank		0.000	04:34
Subject Sample	# 2	0.176	04:35
Breath Volume		1.644	04:35
Air Blank		0.000	04:35
Control Test		0.081	04:36
Air Blank		0.000	04:36
Diagnostics Check		OK	04:36

Data Download Date/Time: 12/20/2019 10:28

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,M,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GELWICKS,W
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:01

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 12/07/2019 05:30:14

Diagnostics Check		OK	05:21	
Air Blank		0.000	05:22	
Control Test		0.083	05:22	
Air Blank		0.000	05:23	
Subject Sample	# 1		05:26	No Sample Provided
Breath Volume		0.000	05:26	
Air Blank		0.000	05:26	
Air Blank		0.000	05:28	
Subject Sample	# 2		05:28	Subject Test Refused
Breath Volume		0.000	05:28	
Air Blank		0.000	05:29	
Control Test		0.081	05:29	
Air Blank		0.000	05:30	
Diagnostics Check		OK	05:30	

Data Download Date/Time: 12/20/2019 10:29

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,M,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HARPIN,R
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:42

SUBJECT'S LAST NAME: RUFFELL

DATE/TIME OF TEST: 12/07/2019 06:09:08

Diagnostics Check		OK	06:02
Air Blank		0.000	06:03
Control Test		0.082	06:03
Air Blank		0.000	06:03
Subject Sample	# 1	0.129	06:04
Breath Volume		1.710	06:04
Air Blank		0.000	06:04
Air Blank		0.000	06:06
Subject Sample	# 2	0.131	06:07
Breath Volume		1.808	06:07
Air Blank		0.000	06:08
Control Test		0.082	06:08
Air Blank		0.000	06:09
Diagnostics Check		OK	06:09

Data Download Date/Time: 12/20/2019 10:29

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 12/07/2019 23:22:33

Diagnostics Check		OK	23:13
Air Blank		0.000	23:14
Control Test		0.082	23:14
Air Blank		0.000	23:15
Subject Sample	# 1	0.149	23:18
Breath Volume		1.699	23:18
Air Blank		0.000	23:18
Air Blank		0.000	23:20
Subject Sample	# 2	0.148	23:20
Breath Volume		2.105	23:20
Air Blank		0.000	23:21
Control Test		0.081	23:21
Air Blank		0.000	23:22
Diagnostics Check		OK	23:22

Data Download Date/Time: 12/20/2019 10:29

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: ANASTACIO PANZO

DATE/TIME OF TEST: 12/08/2019 03:46:25

Diagnostics Check		OK	03:39
Air Blank		0.000	03:39
Control Test		0.081	03:40
Air Blank		0.000	03:40
Subject Sample	# 1	0.149	03:41
Breath Volume		3.675	03:41
Air Blank		0.000	03:42
Air Blank		0.000	03:43
Subject Sample	# 2	0.138	03:44
Breath Volume		1.199	03:44
Air Blank		0.000	03:45
Control Test		0.082	03:45
Air Blank		0.000	03:46
Diagnostics Check		OK	03:46

Data Download Date/Time: 12/20/2019 10:29

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 12/08/2019 04:43:24

Diagnostics Check		OK	04:40	
Air Blank		0.000	04:40	
Control Test		0.082	04:41	
Air Blank		0.000	04:41	
Subject Sample	# 1		04:42	Subject Test Refused
Breath Volume		0.000	04:42	
Air Blank		0.000	04:42	
Control Test		0.081	04:42	
Air Blank		0.000	04:43	
Diagnostics Check		OK	04:43	

Data Download Date/Time: 12/20/2019 10:29

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PEREZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:26

SUBJECT'S LAST NAME: COUSSENS

DATE/TIME OF TEST: 12/08/2019 05:50:01

Diagnostics Check		OK	05:47	
Air Blank		0.000	05:47	
Control Test		0.081	05:47	
Air Blank		0.000	05:48	
Subject Sample	# 1		05:48	Subject Test Refused
Breath Volume		0.000	05:48	
Air Blank		0.000	05:49	
Control Test		0.081	05:49	
Air Blank		0.000	05:49	
Diagnostics Check		OK	05:50	

Data Download Date/Time: 12/20/2019 10:29

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVERETT,I,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TIMMONS,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:18

SUBJECT'S LAST NAME: PARRISH

DATE/TIME OF TEST: 12/08/2019 07:41:05

Diagnostics Check		OK	07:38	
Air Blank		0.000	07:38	
Control Test		0.083	07:38	
Air Blank		0.000	07:39	
Subject Sample	# 1		07:39	Subject Test Refused
Breath Volume		0.000	07:39	
Air Blank		0.000	07:40	
Control Test		0.082	07:40	
Air Blank		0.000	07:41	
Diagnostics Check		OK	07:41	

Data Download Date/Time: 12/20/2019 10:29

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVERETT,I,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DUKIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 06:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:50

SUBJECT'S LAST NAME: MALLORGA VALLES

DATE/TIME OF TEST: 12/08/2019 09:13:57

Diagnostics Check		OK	09:11	
Air Blank		0.000	09:11	
Control Test		0.082	09:11	
Air Blank		0.000	09:12	
Subject Sample	# 1		09:12	Subject Test Refused
Breath Volume		0.000	09:12	
Air Blank		0.000	09:13	
Control Test		0.082	09:13	
Air Blank		0.000	09:13	
Diagnostics Check		OK	09:13	

Data Download Date/Time: 12/20/2019 10:29

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DASHER,JOSHUA,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: BEARD

DATE/TIME OF TEST: 12/08/2019 19:19:17

Diagnostics Check		OK	19:16	
Air Blank		0.000	19:16	
Control Test		0.082	19:17	
Air Blank		0.000	19:17	
Subject Sample	# 1		19:17	Subject Test Refused
Breath Volume		0.000	19:17	
Air Blank		0.000	19:18	
Control Test		0.082	19:18	
Air Blank		0.000	19:19	
Diagnostics Check		OK	19:19	

Data Download Date/Time: 12/20/2019 10:29

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DASHER,JOSHUA,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNS,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:23

SUBJECT'S LAST NAME: SHAW

DATE/TIME OF TEST: 12/08/2019 22:51:20

Diagnostics Check		OK	22:43
Air Blank		0.000	22:44
Control Test		0.081	22:44
Air Blank		0.000	22:45
Subject Sample	# 1	0.063	22:46
Breath Volume		1.156	22:46
Air Blank		0.000	22:47
Air Blank		0.000	22:49
Subject Sample	# 2	0.067	22:49
Breath Volume		1.195	22:49
Air Blank		0.000	22:50
Control Test		0.081	22:50
Air Blank		0.000	22:51
Diagnostics Check		OK	22:51

Data Download Date/Time: 12/20/2019 10:29

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOHN,A,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:12

SUBJECT'S LAST NAME: SEWARD

DATE/TIME OF TEST: 12/09/2019 18:38:52

Diagnostics Check		OK	18:32
Air Blank		0.000	18:33
Control Test		0.081	18:33
Air Blank		0.000	18:34
Subject Sample	# 1	0.011	18:34
Breath Volume		2.089	18:34
Air Blank		0.000	18:35
Air Blank		0.000	18:37
Subject Sample	# 2	0.011	18:37
Breath Volume		2.175	18:37
Air Blank		0.000	18:37
Control Test		0.081	18:38
Air Blank		0.000	18:38
Diagnostics Check		OK	18:38

Data Download Date/Time: 12/20/2019 10:29

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: PESSONEY

DATE/TIME OF TEST: 12/10/2019 04:14:18

Diagnostics Check		OK	04:07
Air Blank		0.000	04:08
Control Test		0.082	04:08
Air Blank		0.000	04:08
Subject Sample	# 1	0.195	04:09
Breath Volume		1.851	04:09
Air Blank		0.000	04:10
Air Blank		0.000	04:11
Subject Sample	# 2	0.215	04:12
Breath Volume		2.867	04:12
Air Blank		0.000	04:13
Control Test		0.081	04:13
Air Blank		0.000	04:14
Diagnostics Check		OK	04:14

Data Download Date/Time: 12/20/2019 10:29

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DASHER,JOSHUA,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:23

SUBJECT'S LAST NAME: LONERGAN

DATE/TIME OF TEST: 12/11/2019 21:51:18

Diagnostics Check		OK	21:43
Air Blank		0.000	21:43
Control Test		0.082	21:43
Air Blank		0.000	21:44
Subject Sample	# 1	0.098	21:45
Breath Volume		1.210	21:45
Air Blank		0.000	21:45
Air Blank		0.000	21:47
Subject Sample	# 2	0.098	21:49
Breath Volume		2.941	21:49
Air Blank		0.000	21:50
Control Test		0.082	21:50
Air Blank		0.000	21:51
Diagnostics Check		OK	21:51

Data Download Date/Time: 12/20/2019 10:29

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DASHER,JOSHUA,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 12/12/2019 03:12:06

Diagnostics Check		OK	03:05
Air Blank		0.000	03:06
Control Test		0.082	03:06
Air Blank		0.000	03:07
Subject Sample	# 1	0.053	03:07
Breath Volume		3.128	03:07
Air Blank		0.000	03:08
Air Blank		0.000	03:10
Subject Sample	# 2	0.054	03:10
Breath Volume		2.726	03:10
Air Blank		0.000	03:11
Control Test		0.083	03:11
Air Blank		0.000	03:12
Diagnostics Check		OK	03:12

Data Download Date/Time: 12/20/2019 10:29

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 12/13/2019 04:15:54

Diagnostics Check		OK	04:09
Air Blank		0.000	04:10
Control Test		0.081	04:10
Air Blank		0.000	04:11
Subject Sample	# 1	0.105	04:11
Breath Volume		1.792	04:11
Air Blank		0.000	04:12
Air Blank		0.000	04:13
Subject Sample	# 2	0.105	04:14
Breath Volume		1.886	04:14
Air Blank		0.000	04:14
Control Test		0.080	04:15
Air Blank		0.000	04:15
Diagnostics Check		OK	04:15

Data Download Date/Time: 12/20/2019 10:29

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,M,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HARPIN,R
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: FIGURA

DATE/TIME OF TEST: 12/14/2019 01:41:27

Diagnostics Check		OK	01:35
Air Blank		0.000	01:35
Control Test		0.081	01:35
Air Blank		0.000	01:36
Subject Sample	# 1	0.045	01:36
Breath Volume		1.496	01:36
Air Blank		0.000	01:37
Air Blank		0.000	01:39
Subject Sample	# 2	0.043	01:39
Breath Volume		1.628	01:39
Air Blank		0.000	01:40
Control Test		0.081	01:40
Air Blank		0.000	01:41
Diagnostics Check		OK	01:41

Data Download Date/Time: 12/20/2019 10:29

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,M,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: ROBLERO DIAZ

DATE/TIME OF TEST: 12/16/2019 03:57:55

Diagnostics Check		OK	03:49
Air Blank		0.000	03:49
Control Test		0.081	03:50
Air Blank		0.000	03:50
Subject Sample	# 1	0.222	03:52
Breath Volume		2.113	03:52
Air Blank		0.000	03:53
Air Blank		0.000	03:55
Subject Sample	# 2	0.226	03:56
Breath Volume		2.230	03:56
Air Blank		0.000	03:56
Control Test		0.081	03:57
Air Blank		0.000	03:57
Diagnostics Check		OK	03:57

Data Download Date/Time: 12/20/2019 10:29

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: JAMES

DATE/TIME OF TEST: 12/16/2019 23:24:57

Diagnostics Check		OK	23:18
Air Blank		0.000	23:18
Control Test		0.081	23:19
Air Blank		0.000	23:19
Subject Sample	# 1	0.119	23:20
Breath Volume		2.437	23:20
Air Blank		0.000	23:21
Air Blank		0.000	23:22
Subject Sample	# 2	0.121	23:23
Breath Volume		2.921	23:23
Air Blank		0.000	23:24
Control Test		0.080	23:24
Air Blank		0.000	23:24
Diagnostics Check		OK	23:24

Data Download Date/Time: 12/20/2019 10:29

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 12/17/2019 04:17:45

Diagnostics Check		OK	04:10
Air Blank		0.000	04:11
Control Test		0.081	04:11
Air Blank		0.000	04:11
Subject Sample	# 1	0.000	04:12
Breath Volume		1.250	04:12
Air Blank		0.000	04:13
Air Blank		0.000	04:15
Subject Sample	# 2	0.000	04:16
Breath Volume		1.855	04:16
Air Blank		0.000	04:16
Control Test		0.081	04:17
Air Blank		0.000	04:17
Diagnostics Check		OK	04:17

Data Download Date/Time: 12/20/2019 10:29

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAMONT,J
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:02

SUBJECT'S LAST NAME: MALDONADO

DATE/TIME OF TEST: 12/17/2019 06:25:41

Diagnostics Check		OK	06:22	
Air Blank		0.000	06:23	
Control Test		0.080	06:23	
Air Blank		0.000	06:24	
Subject Sample	# 1		06:24	Subject Test Refused
Breath Volume		0.000	06:24	
Air Blank		0.000	06:24	
Control Test		0.080	06:25	
Air Blank		0.000	06:25	
Diagnostics Check		OK	06:25	

Data Download Date/Time: 12/20/2019 10:29

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVERETT,I,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ANDRES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 07:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:55

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 12/17/2019 09:18:08

Diagnostics Check		OK	09:15	
Air Blank		0.000	09:15	
Control Test		0.081	09:16	
Air Blank		0.000	09:16	
Subject Sample	# 1		09:16	Subject Test Refused
Breath Volume		0.000	09:16	
Air Blank		0.000	09:17	
Control Test		0.080	09:17	
Air Blank		0.000	09:18	
Diagnostics Check		OK	09:18	

Data Download Date/Time: 12/20/2019 10:29

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DASHER,JOSHUA,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: GULLETT

DATE/TIME OF TEST: 12/18/2019 00:47:25

Diagnostics Check		OK	00:41
Air Blank		0.000	00:41
Control Test		0.082	00:42
Air Blank		0.000	00:42
Subject Sample	# 1	0.087	00:43
Breath Volume		2.367	00:43
Air Blank		0.000	00:43
Air Blank		0.000	00:45
Subject Sample	# 2	0.085	00:45
Breath Volume		2.558	00:45
Air Blank		0.000	00:46
Control Test		0.082	00:46
Air Blank		0.000	00:47
Diagnostics Check		OK	00:47

Data Download Date/Time: 12/20/2019 10:29

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DASHER,JOSHUA,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: PUCCIO

DATE/TIME OF TEST: 12/18/2019 02:52:28

Diagnostics Check		OK	02:49	
Air Blank		0.000	02:50	
Control Test		0.082	02:50	
Air Blank		0.000	02:50	
Subject Sample	# 1		02:51	Subject Test Refused
Breath Volume		0.000	02:51	
Air Blank		0.000	02:51	
Control Test		0.082	02:51	
Air Blank		0.000	02:52	
Diagnostics Check		OK	02:52	

Data Download Date/Time: 12/20/2019 10:29

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: KOCHER

DATE/TIME OF TEST: 12/19/2019 00:10:31

Diagnostics Check		OK	00:04
Air Blank		0.000	00:04
Control Test		0.083	00:04
Air Blank		0.000	00:05
Subject Sample	# 1	0.149	00:05
Breath Volume		1.910	00:05
Air Blank		0.000	00:06
Air Blank		0.000	00:08
Subject Sample	# 2	0.141	00:08
Breath Volume		1.613	00:08
Air Blank		0.000	00:09
Control Test		0.082	00:09
Air Blank		0.000	00:10
Diagnostics Check		OK	00:10

Data Download Date/Time: 12/20/2019 10:29

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,M,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: GRAHAM

DATE/TIME OF TEST: 12/21/2019 23:28:45

Diagnostics Check		OK	23:25	
Air Blank		0.000	23:26	
Control Test		0.082	23:26	
Air Blank		0.000	23:27	
Subject Sample	# 1		23:27	Subject Test Refused
Breath Volume		0.000	23:27	
Air Blank		0.000	23:27	
Control Test		0.081	23:28	
Air Blank		0.000	23:28	
Diagnostics Check		OK	23:28	

Data Download Date/Time: 01/16/2020 07:26

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,M,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: VELASQUEZ

DATE/TIME OF TEST: 12/22/2019 04:55:02

Diagnostics Check		OK	04:47
Air Blank		0.000	04:48
Control Test		0.082	04:48
Air Blank		0.000	04:49
Subject Sample	# 1	0.119	04:49
Breath Volume		1.210	04:49
Air Blank		0.000	04:50
Air Blank		0.000	04:52
Subject Sample	# 2	0.114	04:53
Breath Volume		1.886	04:53
Air Blank		0.000	04:54
Control Test		0.080	04:54
Air Blank		0.000	04:54
Diagnostics Check		OK	04:55

Data Download Date/Time: 01/16/2020 07:26

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,M,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CROSS,W
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:59

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 12/22/2019 06:25:59

Diagnostics Check		OK	06:19
Air Blank		0.000	06:20
Control Test		0.080	06:20
Air Blank		0.000	06:20
Subject Sample	# 1	0.019	06:21
Breath Volume		1.394	06:21
Air Blank		0.000	06:22
Air Blank		0.000	06:23
Subject Sample	# 2	0.018	06:24
Breath Volume		1.531	06:24
Air Blank		0.000	06:25
Control Test		0.080	06:25
Air Blank		0.000	06:25
Diagnostics Check		OK	06:25

Data Download Date/Time: 01/16/2020 07:26

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: WOODWARD

DATE/TIME OF TEST: 12/22/2019 23:17:11

Diagnostics Check		OK	23:14	
Air Blank		0.000	23:14	
Control Test		0.080	23:14	
Air Blank		0.000	23:15	
Subject Sample	# 1		23:15	Subject Test Refused
Breath Volume		0.000	23:15	
Air Blank		0.000	23:16	
Control Test		0.081	23:16	
Air Blank		0.000	23:17	
Diagnostics Check		OK	23:17	

Data Download Date/Time: 01/16/2020 07:26

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: ROOT

DATE/TIME OF TEST: 12/24/2019 02:24:43

Diagnostics Check		OK	02:18
Air Blank		0.000	02:19
Control Test		0.082	02:19
Air Blank		0.000	02:19
Subject Sample	# 1	0.209	02:20
Breath Volume		4.035	02:20
Air Blank		0.000	02:20
Air Blank		0.000	02:22
Subject Sample	# 2	0.207	02:23
Breath Volume		3.023	02:23
Air Blank		0.000	02:23
Control Test		0.081	02:24
Air Blank		0.000	02:24
Diagnostics Check		OK	02:24

Data Download Date/Time: 01/16/2020 07:26

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: SQUAIRE

DATE/TIME OF TEST: 12/24/2019 23:09:04

Diagnostics Check		OK	23:02
Air Blank		0.000	23:02
Control Test		0.081	23:02
Air Blank		0.000	23:03
Subject Sample	# 1	0.152	23:04
Breath Volume		1.292	23:04
Air Blank		0.000	23:05
Air Blank		0.000	23:06
Subject Sample	# 2	0.161	23:07
Breath Volume		1.781	23:07
Air Blank		0.000	23:08
Control Test		0.080	23:08
Air Blank		0.000	23:08
Diagnostics Check		OK	23:09

Data Download Date/Time: 01/16/2020 07:26

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,M,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: LARMOYEUX

DATE/TIME OF TEST: 12/25/2019 01:25:20

Diagnostics Check		OK	01:22	
Air Blank		0.000	01:22	
Control Test		0.081	01:23	
Air Blank		0.000	01:23	
Subject Sample	# 1		01:23	Subject Test Refused
Breath Volume		0.000	01:23	
Air Blank		0.000	01:24	
Control Test		0.082	01:24	
Air Blank		0.000	01:25	
Diagnostics Check		OK	01:25	

Data Download Date/Time: 01/16/2020 07:26

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: FULGHUM

DATE/TIME OF TEST: 12/25/2019 02:50:15

Diagnostics Check		OK	02:44
Air Blank		0.000	02:44
Control Test		0.080	02:44
Air Blank		0.000	02:45
Subject Sample	# 1	0.135	02:45
Breath Volume		2.191	02:45
Air Blank		0.000	02:46
Air Blank		0.000	02:48
Subject Sample	# 2	0.135	02:48
Breath Volume		2.101	02:48
Air Blank		0.000	02:49
Control Test		0.081	02:49
Air Blank		0.000	02:50
Diagnostics Check		OK	02:50

Data Download Date/Time: 01/16/2020 07:26

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:03

SUBJECT'S LAST NAME: BROY

DATE/TIME OF TEST: 12/25/2019 21:26:40

Diagnostics Check		OK	21:23	
Air Blank		0.000	21:24	
Control Test		0.082	21:24	
Air Blank		0.000	21:24	
Subject Sample	# 1		21:25	Subject Test Refused
Breath Volume		0.000	21:25	
Air Blank		0.000	21:25	
Control Test		0.081	21:26	
Air Blank		0.000	21:26	
Diagnostics Check		OK	21:26	

Data Download Date/Time: 01/16/2020 07:26

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,M,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FILBERT,E
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 12/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: BAUMAN

DATE/TIME OF TEST: 12/25/2019 23:40:14

Diagnostics Check		OK	23:37	
Air Blank		0.000	23:37	
Control Test		0.082	23:38	
Air Blank		0.000	23:38	
Subject Sample	# 1		23:38	Subject Test Refused
Breath Volume		0.000	23:38	
Air Blank		0.000	23:39	
Control Test		0.081	23:39	
Air Blank		0.000	23:40	
Diagnostics Check		OK	23:40	

Data Download Date/Time: 01/16/2020 07:26

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ACKERMAN-CARTER,O
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: MAHONEY

DATE/TIME OF TEST: 12/26/2019 02:37:47

Diagnostics Check		OK	02:31
Air Blank		0.000	02:31
Control Test		0.081	02:32
Air Blank		0.000	02:32
Subject Sample	# 1	0.185	02:33
Breath Volume		2.121	02:33
Air Blank		0.000	02:33
Air Blank		0.000	02:35
Subject Sample	# 2	0.186	02:36
Breath Volume		2.238	02:36
Air Blank		0.000	02:36
Control Test		0.081	02:37
Air Blank		0.000	02:37
Diagnostics Check		OK	02:37

Data Download Date/Time: 01/16/2020 07:26

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DASHER,JOSHUA,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TORRES,W
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 12/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:02

SUBJECT'S LAST NAME: MCCUE

DATE/TIME OF TEST: 12/26/2019 19:29:01

Diagnostics Check		OK	19:22
Air Blank		0.000	19:23
Control Test		0.081	19:23
Air Blank		0.000	19:23
Subject Sample	# 1	0.204	19:24
Breath Volume		2.769	19:24
Air Blank		0.000	19:25
Air Blank		0.000	19:26
Subject Sample	# 2	0.200	19:27
Breath Volume		2.687	19:27
Air Blank		0.000	19:28
Control Test		0.080	19:28
Air Blank		0.000	19:28
Diagnostics Check		OK	19:29

Data Download Date/Time: 01/16/2020 07:26

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 12/27/2019 01:12:56

Diagnostics Check		OK	01:06
Air Blank		0.000	01:07
Control Test		0.082	01:07
Air Blank		0.000	01:08
Subject Sample	# 1	0.166	01:08
Breath Volume		2.800	01:08
Air Blank		0.000	01:09
Air Blank		0.000	01:10
Subject Sample	# 2	0.168	01:11
Breath Volume		3.058	01:11
Air Blank		0.000	01:12
Control Test		0.081	01:12
Air Blank		0.000	01:12
Diagnostics Check		OK	01:12

Data Download Date/Time: 01/16/2020 07:26

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DASHER,JOSHUA,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNS,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: WARF

DATE/TIME OF TEST: 12/27/2019 03:33:12

Diagnostics Check		OK	03:26
Air Blank		0.000	03:26
Control Test		0.081	03:26
Air Blank		0.000	03:27
Subject Sample	# 1	0.117	03:27
Breath Volume		1.140	03:27
Air Blank		0.000	03:28
Air Blank		0.000	03:30
Subject Sample	# 2	0.120	03:31
Breath Volume		1.171	03:31
Air Blank		0.000	03:32
Control Test		0.081	03:32
Air Blank		0.000	03:33
Diagnostics Check		OK	03:33

Data Download Date/Time: 01/16/2020 07:26

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DASHER,JOSHUA,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: JUBANI

DATE/TIME OF TEST: 12/27/2019 23:52:57

Diagnostics Check		OK	23:46
Air Blank		0.000	23:47
Control Test		0.081	23:47
Air Blank		0.000	23:48
Subject Sample	# 1	0.000	23:48
Breath Volume		2.992	23:48
Air Blank		0.000	23:49
Air Blank		0.000	23:51
Subject Sample	# 2	0.000	23:51
Breath Volume		2.867	23:51
Air Blank		0.000	23:52
Control Test		0.081	23:52
Air Blank		0.000	23:52
Diagnostics Check		OK	23:52

Data Download Date/Time: 01/16/2020 07:26

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DASHER,JOSHUA,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 12/28/2019 04:46:05

Diagnostics Check		OK	04:40
Air Blank		0.000	04:40
Control Test		0.082	04:40
Air Blank		0.000	04:41
Subject Sample	# 1	0.165	04:41
Breath Volume		1.742	04:41
Air Blank		0.000	04:42
Air Blank		0.000	04:44
Subject Sample	# 2	0.162	04:44
Breath Volume		1.792	04:44
Air Blank		0.000	04:45
Control Test		0.081	04:45
Air Blank		0.000	04:46
Diagnostics Check		OK	04:46

Data Download Date/Time: 01/16/2020 07:26

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOHN,A,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 08:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:34

SUBJECT'S LAST NAME: BUTTERFIELD

DATE/TIME OF TEST: 12/28/2019 10:01:23

Diagnostics Check		OK	09:55
Air Blank		0.000	09:55
Control Test		0.081	09:55
Air Blank		0.000	09:56
Subject Sample	# 1	0.179	09:56
Breath Volume		3.218	09:56
Air Blank		0.000	09:57
Air Blank		0.000	09:59
Subject Sample	# 2	0.185	09:59
Breath Volume		3.281	09:59
Air Blank		0.000	10:00
Control Test		0.081	10:00
Air Blank		0.000	10:01
Diagnostics Check		OK	10:01

Data Download Date/Time: 01/16/2020 07:26

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DASHER,JOSHUA,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 12/28/2019 22:34:28

Diagnostics Check		OK	22:28
Air Blank		0.000	22:28
Control Test		0.082	22:28
Air Blank		0.000	22:29
Subject Sample	# 1	0.126	22:30
Breath Volume		1.656	22:30
Air Blank		0.000	22:30
Air Blank		0.000	22:32
Subject Sample	# 2	0.116	22:32
Breath Volume		1.527	22:32
Air Blank		0.000	22:33
Control Test		0.081	22:33
Air Blank		0.000	22:34
Diagnostics Check		OK	22:34

Data Download Date/Time: 01/16/2020 07:26

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVERETT,I,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SINGLETON,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 11:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:20

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 12/29/2019 13:48:37

Diagnostics Check		OK	13:42
Air Blank		0.000	13:42
Control Test		0.081	13:42
Air Blank		0.000	13:43
Subject Sample	# 1	0.125	13:43
Breath Volume		3.628	13:43
Air Blank		0.000	13:44
Air Blank		0.000	13:46
Subject Sample	# 2	0.124	13:47
Breath Volume		3.625	13:47
Air Blank		0.000	13:47
Control Test		0.081	13:48
Air Blank		0.000	13:48
Diagnostics Check		OK	13:48

Data Download Date/Time: 01/16/2020 07:26

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DASHER,JOSHUA,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNS,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: VAN

DATE/TIME OF TEST: 12/29/2019 23:20:07

Diagnostics Check		OK	23:14
Air Blank		0.000	23:14
Control Test		0.081	23:14
Air Blank		0.000	23:15
Subject Sample	# 1	0.000	23:15
Breath Volume		3.343	23:15
Air Blank		0.000	23:16
Air Blank		0.000	23:18
Subject Sample	# 2	0.000	23:18
Breath Volume		2.761	23:18
Air Blank		0.000	23:19
Control Test		0.082	23:19
Air Blank		0.000	23:20
Diagnostics Check		OK	23:20

Data Download Date/Time: 01/16/2020 07:26

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DASHER,JOSHUA,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: YOUNG,S
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 12/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: PICHARDO MEZA

DATE/TIME OF TEST: 12/30/2019 04:53:27

Diagnostics Check		OK	04:47
Air Blank		0.000	04:47
Control Test		0.081	04:48
Air Blank		0.000	04:48
Subject Sample	# 1	0.134	04:49
Breath Volume		1.734	04:49
Air Blank		0.000	04:49
Air Blank		0.000	04:51
Subject Sample	# 2	0.128	04:51
Breath Volume		1.546	04:51
Air Blank		0.000	04:52
Control Test		0.081	04:52
Air Blank		0.000	04:53
Diagnostics Check		OK	04:53

Data Download Date/Time: 01/16/2020 07:26

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EBERT,L
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:47

SUBJECT'S LAST NAME: GODDARD

DATE/TIME OF TEST: 12/30/2019 20:11:47

Diagnostics Check		OK	20:08	
Air Blank		0.000	20:09	
Control Test		0.082	20:09	
Air Blank		0.000	20:10	
Subject Sample	# 1		20:10	Subject Test Refused
Breath Volume		0.000	20:10	
Air Blank		0.000	20:10	
Control Test		0.082	20:11	
Air Blank		0.000	20:11	
Diagnostics Check		OK	20:11	

Data Download Date/Time: 01/16/2020 07:26

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CURRY,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: HUMPHREY

DATE/TIME OF TEST: 12/31/2019 22:45:35

Diagnostics Check		OK	22:39
Air Blank		0.000	22:39
Control Test		0.082	22:40
Air Blank		0.000	22:40
Subject Sample	# 1	0.181	22:41
Breath Volume		2.933	22:41
Air Blank		0.000	22:41
Air Blank		0.000	22:43
Subject Sample	# 2	0.195	22:43
Breath Volume		3.269	22:43
Air Blank		0.000	22:44
Control Test		0.082	22:44
Air Blank		0.000	22:45
Diagnostics Check		OK	22:45

Data Download Date/Time: 01/16/2020 07:26

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001073

INSTRUMENT REGISTERED TO: DADE CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSENBLOOM,MATTHEW,S
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: DOUGHERTY,THOMAS
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 12/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A3
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 07:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:47

SUBJECT'S LAST NAME: SANTOS ACEVEDO

DATE/TIME OF TEST: 12/24/2019 08:34:38

Diagnostics Check		OK	08:28
Air Blank		0.000	08:28
Control Test		0.079	08:29
Air Blank		0.000	08:29
Subject Sample	# 1	0.000	08:30
Breath Volume		3.324	08:30
Air Blank		0.000	08:30
Air Blank		0.000	08:32
Subject Sample	# 2	0.000	08:33
Breath Volume		3.457	08:33
Air Blank		0.000	08:33
Control Test		0.079	08:34
Air Blank		0.000	08:34
Diagnostics Check		OK	08:34

Data Download Date/Time: 01/15/2020 16:28

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001076

INSTRUMENT REGISTERED TO: ZEPHYRHILLS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAINRIGHT,DAVID,T
BREATH TEST OPERATOR AGENCY: Zephyrhills PD
ARRESTING OFFICER: RUSSELL,JASON
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 12/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: LUGER

DATE/TIME OF TEST: 12/08/2019 01:23:43

Diagnostics Check		OK	01:17
Air Blank		0.000	01:18
Control Test		0.081	01:18
Air Blank		0.000	01:19
Subject Sample	# 1	0.000	01:19
Breath Volume		2.246	01:19
Air Blank		0.000	01:20
Air Blank		0.000	01:22
Subject Sample	# 2	0.000	01:22
Breath Volume		1.898	01:22
Air Blank		0.000	01:22
Control Test		0.081	01:23
Air Blank		0.000	01:23
Diagnostics Check		OK	01:23

Data Download Date/Time: 01/08/2020 16:15

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINES,PAUL,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CRUM,ZACKARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: MCCOMESKEY

DATE/TIME OF TEST: 12/06/2019 02:02:52

Diagnostics Check		OK	01:51	
Air Blank		0.000	01:51	
Control Test		0.079	01:52	
Air Blank		0.000	01:52	
Subject Sample	# 1		01:55	No Sample Provided
Breath Volume		0.000	01:55	
Air Blank		0.000	01:56	
Air Blank		0.000	01:58	
Subject Sample	# 2		02:01	No Sample Provided
Breath Volume		0.000	02:01	
Air Blank		0.000	02:01	
Control Test		0.080	02:02	
Air Blank		0.000	02:02	
Diagnostics Check		OK	02:02	

Data Download Date/Time: 12/23/2019 09:53

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCULLOUGH,SEAN,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LESPERANCE,DANIEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: HOCK

DATE/TIME OF TEST: 12/07/2019 00:50:08

Diagnostics Check		OK	00:43
Air Blank		0.000	00:44
Control Test		0.078	00:44
Air Blank		0.000	00:45
Subject Sample	# 1	0.304	00:45
Breath Volume		4.371	00:45
Air Blank		0.000	00:46
Air Blank		0.000	00:48
Subject Sample	# 2	0.299	00:48
Breath Volume		4.406	00:48
Air Blank		0.000	00:49
Control Test		0.078	00:49
Air Blank		0.000	00:50
Diagnostics Check		OK	00:50

Data Download Date/Time: 12/23/2019 09:53

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOCKERS,M,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CLARK,B
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 09:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:33

SUBJECT'S LAST NAME: GUEL

DATE/TIME OF TEST: 12/17/2019 09:59:47

Diagnostics Check		OK	09:53
Air Blank		0.000	09:54
Control Test		0.080	09:54
Air Blank		0.000	09:54
Subject Sample	# 1	0.000	09:55
Breath Volume		2.222	09:55
Air Blank		0.000	09:55
Air Blank		0.000	09:57
Subject Sample	# 2	0.000	09:58
Breath Volume		1.601	09:58
Air Blank		0.000	09:58
Control Test		0.079	09:59
Air Blank		0.000	09:59
Diagnostics Check		OK	09:59

Data Download Date/Time: 12/23/2019 09:53

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUM,ZACKARY,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HINES,Z
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:39

SUBJECT'S LAST NAME: MORROW

DATE/TIME OF TEST: 12/17/2019 19:08:12

Diagnostics Check		OK	18:59	
Air Blank		0.000	19:00	
Control Test		0.080	19:00	
Air Blank		0.000	19:00	
Subject Sample	# 1		19:04	No Sample Provided
Breath Volume		0.000	19:04	
Air Blank		0.000	19:04	
Air Blank		0.000	19:06	
Subject Sample	# 2		19:06	Subject Test Refused
Breath Volume		0.000	19:06	
Air Blank		0.000	19:07	
Control Test		0.079	19:07	
Air Blank		0.000	19:08	
Diagnostics Check		OK	19:08	

Data Download Date/Time: 12/23/2019 09:53

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINES,PAUL,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LAMOUR,ZACHARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: LENTSCH

DATE/TIME OF TEST: 12/17/2019 21:48:38

Diagnostics Check		OK	21:42
Air Blank		0.000	21:42
Control Test		0.079	21:43
Air Blank		0.000	21:43
Subject Sample	# 1	0.228	21:44
Breath Volume		1.167	21:44
Air Blank		0.000	21:44
Air Blank		0.000	21:46
Subject Sample	# 2	0.233	21:47
Breath Volume		1.800	21:47
Air Blank		0.000	21:47
Control Test		0.079	21:48
Air Blank		0.000	21:48
Diagnostics Check		OK	21:48

Data Download Date/Time: 12/23/2019 09:53

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINES,PAUL,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCCULLOUGH,SEAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: GUNSAULLUS

DATE/TIME OF TEST: 12/19/2019 01:13:49

Diagnostics Check		OK	01:07
Air Blank		0.000	01:07
Control Test		0.078	01:08
Air Blank		0.000	01:08
Subject Sample	# 1	0.074	01:09
Breath Volume		1.621	01:09
Air Blank		0.000	01:09
Air Blank		0.000	01:11
Subject Sample	# 2	0.074	01:12
Breath Volume		3.945	01:12
Air Blank		0.000	01:12
Control Test		0.081	01:13
Air Blank		0.000	01:13
Diagnostics Check		OK	01:13

Data Download Date/Time: 12/23/2019 09:53

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINES,PAUL,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: POST,ALEXANDER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: FRIEL

DATE/TIME OF TEST: 12/19/2019 02:05:40

Diagnostics Check		OK	01:56
Air Blank		0.000	01:57
Control Test		0.080	01:57
Air Blank		0.000	01:58
Subject Sample	# 1	0.167	02:00
Breath Volume		2.046	02:00
Air Blank		0.000	02:01
Air Blank		0.000	02:03
Subject Sample	# 2	0.164	02:04
Breath Volume		2.605	02:04
Air Blank		0.000	02:04
Control Test		0.080	02:05
Air Blank		0.000	02:05
Diagnostics Check		OK	02:05

Data Download Date/Time: 12/23/2019 09:53

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCULLOUGH,SEAN,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LAMOUR,ZACHARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: GARIB

DATE/TIME OF TEST: 12/20/2019 02:38:53

Diagnostics Check		OK	02:32
Air Blank		0.000	02:33
Control Test		0.080	02:33
Air Blank		0.000	02:33
Subject Sample	# 1	0.121	02:34
Breath Volume		3.816	02:34
Air Blank		0.000	02:34
Air Blank		0.000	02:36
Subject Sample	# 2	0.125	02:37
Breath Volume		3.660	02:37
Air Blank		0.000	02:37
Control Test		0.079	02:38
Air Blank		0.000	02:38
Diagnostics Check		OK	02:38

Data Download Date/Time: 12/23/2019 09:53

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAIA,MICHAEL,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: GODERICH,KRISTINA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:16

SUBJECT'S LAST NAME: RIOS RAMOS

DATE/TIME OF TEST: 12/22/2019 06:03:40

Diagnostics Check		OK	05:56
Air Blank		0.000	05:57
Control Test		0.078	05:57
Air Blank		0.000	05:57
Subject Sample	# 1	0.159	05:58
Breath Volume		2.218	05:58
Air Blank		0.000	05:59
Air Blank		0.000	06:01
Subject Sample	# 2	0.165	06:02
Breath Volume		3.320	06:02
Air Blank		0.000	06:02
Control Test		0.077	06:03
Air Blank		0.000	06:03
Diagnostics Check		OK	06:03

Data Download Date/Time: 12/23/2019 09:53

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRNKA,L,T
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: NEVERLINE,D
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: HARKINS II

DATE/TIME OF TEST: 12/02/2019 19:57:40

Diagnostics Check		OK	19:51
Air Blank		0.000	19:51
Control Test		0.081	19:52
Air Blank		0.000	19:52
Subject Sample	# 1	0.232	19:53
Breath Volume		2.503	19:53
Air Blank		0.000	19:53
Air Blank		0.000	19:55
Subject Sample	# 2	0.231	19:56
Breath Volume		2.910	19:56
Air Blank		0.000	19:56
Control Test		0.081	19:57
Air Blank		0.000	19:57
Diagnostics Check		OK	19:57

Data Download Date/Time: 12/05/2019 11:22

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASTINGS,KENNETH,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: HASTINGS,KENNETH
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: KELSEY

DATE/TIME OF TEST: 12/05/2019 01:25:17

Diagnostics Check		OK	01:22	
Air Blank		0.000	01:22	
Control Test		0.081	01:23	
Air Blank		0.000	01:23	
Subject Sample	# 1		01:23	Subject Test Refused
Breath Volume		0.000	01:23	
Air Blank		0.000	01:24	
Control Test		0.080	01:24	
Air Blank		0.000	01:25	
Diagnostics Check		OK	01:25	

Data Download Date/Time: 12/05/2019 11:22

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASTINGS,KENNETH,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: HASTINGS,KENNETH
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: NARAIN

DATE/TIME OF TEST: 12/08/2019 02:34:01

Diagnostics Check		OK	02:25
Air Blank		0.000	02:26
Control Test		0.080	02:26
Air Blank		0.000	02:26
Subject Sample	# 1	0.211	02:29
Breath Volume		1.308	02:29
Air Blank		0.000	02:29
Air Blank		0.000	02:31
Subject Sample	# 2	0.224	02:32
Breath Volume		1.429	02:32
Air Blank		0.000	02:33
Control Test		0.080	02:33
Air Blank		0.000	02:33
Diagnostics Check		OK	02:34

Data Download Date/Time: 12/18/2019 11:10

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 12/14/2019 00:28:42

Diagnostics Check		OK	00:25	
Air Blank		0.000	00:26	
Control Test		0.080	00:26	
Air Blank		0.000	00:27	
Subject Sample	# 1		00:27	Subject Test Refused
Breath Volume		0.000	00:27	
Air Blank		0.000	00:27	
Control Test		0.081	00:28	
Air Blank		0.000	00:28	
Diagnostics Check		OK	00:28	

Data Download Date/Time: 12/18/2019 11:10

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASTINGS,KENNETH,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: HASTINGS,KENNETH
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: LAKILAK

DATE/TIME OF TEST: 12/14/2019 03:42:57

Diagnostics Check		OK	03:36
Air Blank		0.000	03:37
Control Test		0.080	03:37
Air Blank		0.000	03:38
Subject Sample	# 1	0.118	03:38
Breath Volume		3.019	03:38
Air Blank		0.000	03:39
Air Blank		0.000	03:41
Subject Sample	# 2	0.121	03:41
Breath Volume		2.292	03:41
Air Blank		0.000	03:42
Control Test		0.080	03:42
Air Blank		0.000	03:42
Diagnostics Check		OK	03:42

Data Download Date/Time: 12/18/2019 11:10

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASTINGS,KENNETH,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: WARD,LOGAN
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: GARCIA-MARIN

DATE/TIME OF TEST: 12/14/2019 04:21:52

Diagnostics Check		OK	04:15
Air Blank		0.000	04:15
Control Test		0.080	04:16
Air Blank		0.000	04:16
Subject Sample	# 1	0.138	04:17
Breath Volume		2.500	04:17
Air Blank		0.000	04:17
Air Blank		0.000	04:19
Subject Sample	# 2	0.138	04:20
Breath Volume		1.253	04:20
Air Blank		0.000	04:20
Control Test		0.080	04:21
Air Blank		0.000	04:21
Diagnostics Check		OK	04:21

Data Download Date/Time: 12/18/2019 11:10

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASTINGS,KENNETH,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: HASTINGS,KENNETH
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: GRATTIDGE

DATE/TIME OF TEST: 12/15/2019 02:23:40

Diagnostics Check		OK	02:17
Air Blank		0.000	02:18
Control Test		0.080	02:18
Air Blank		0.000	02:18
Subject Sample	# 1	0.246	02:19
Breath Volume		1.425	02:19
Air Blank		0.000	02:19
Air Blank		0.000	02:21
Subject Sample	# 2	0.262	02:22
Breath Volume		4.417	02:22
Air Blank		0.000	02:22
Control Test		0.081	02:23
Air Blank		0.000	02:23
Diagnostics Check		OK	02:23

Data Download Date/Time: 12/18/2019 11:10

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KATZ,JUSTIN,T
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: KATZ,JUSTIN
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:06

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 12/15/2019 15:30:32

Diagnostics Check		OK	15:26	
Air Blank		0.000	15:27	
Control Test		0.081	15:27	
Air Blank		0.000	15:28	
Subject Sample	# 1		15:29	Subject Test Refused
Breath Volume		0.000	15:29	
Air Blank		0.000	15:29	
Control Test		0.080	15:29	
Air Blank		0.000	15:30	
Diagnostics Check		OK	15:30	

Data Download Date/Time: 12/18/2019 11:10

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOSIER,MATTHEW,D
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: GOSIER,MATTHEW
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: GOMEZ BORIA

DATE/TIME OF TEST: 12/16/2019 03:20:03

Diagnostics Check		OK	03:11
Air Blank		0.000	03:11
Control Test		0.080	03:12
Air Blank		0.000	03:12
Subject Sample	# 1	0.191	03:13
Breath Volume		1.371	03:13
Air Blank		0.000	03:14
Air Blank		0.000	03:15
Subject Sample	# 2	0.179	03:18
Breath Volume		1.308	03:18
Air Blank		0.000	03:19
Control Test		0.080	03:19
Air Blank		0.000	03:19
Diagnostics Check		OK	03:20

Data Download Date/Time: 12/18/2019 11:10

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASTINGS,KENNETH,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: URBANIAK,JEFFREY
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 12/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: SHAFFER

DATE/TIME OF TEST: 12/19/2019 23:45:12

Diagnostics Check		OK	23:39
Air Blank		0.000	23:39
Control Test		0.081	23:39
Air Blank		0.000	23:40
Subject Sample	# 1	0.186	23:40
Breath Volume		2.582	23:40
Air Blank		0.000	23:41
Air Blank		0.000	23:43
Subject Sample	# 2	0.185	23:43
Breath Volume		2.582	23:43
Air Blank		0.000	23:44
Control Test		0.081	23:44
Air Blank		0.000	23:45
Diagnostics Check		OK	23:45

Data Download Date/Time: 01/16/2020 15:19

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 12/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:14

SUBJECT'S LAST NAME: SCHNELZER

DATE/TIME OF TEST: 12/21/2019 19:45:33

Diagnostics Check		OK	19:36	
Air Blank		0.000	19:37	
Control Test		0.081	19:37	
Air Blank		0.000	19:38	
Subject Sample	# 1		19:41	No Sample Provided
Breath Volume		0.000	19:41	
Air Blank		0.000	19:41	
Air Blank		0.000	19:43	
Subject Sample	# 2		19:44	Subject Test Refused
Breath Volume		0.000	19:44	
Air Blank		0.000	19:44	
Control Test		0.080	19:44	
Air Blank		0.000	19:45	
Diagnostics Check		OK	19:45	

Data Download Date/Time: 01/16/2020 15:19

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASTINGS,KENNETH,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: HASTINGS,KENNETH
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 12/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:49

SUBJECT'S LAST NAME: SCURRIA

DATE/TIME OF TEST: 12/25/2019 21:17:29

Diagnostics Check		OK	21:10
Air Blank		0.000	21:11
Control Test		0.080	21:11
Air Blank		0.000	21:11
Subject Sample	# 1	0.000	21:13
Breath Volume		1.660	21:13
Air Blank		0.000	21:13
Air Blank		0.000	21:15
Subject Sample	# 2	0.000	21:16
Breath Volume		1.656	21:16
Air Blank		0.000	21:16
Control Test		0.081	21:16
Air Blank		0.000	21:17
Diagnostics Check		OK	21:17

Data Download Date/Time: 01/16/2020 15:19

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASTINGS,KENNETH,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: HASTINGS,KENNETH
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 12/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: DUDLEY

DATE/TIME OF TEST: 12/27/2019 00:11:50

Diagnostics Check		OK	00:08	
Air Blank		0.000	00:09	
Control Test		0.081	00:09	
Air Blank		0.000	00:10	
Subject Sample	# 1		00:10	Subject Test Refused
Breath Volume		0.000	00:10	
Air Blank		0.000	00:10	
Control Test		0.080	00:11	
Air Blank		0.000	00:11	
Diagnostics Check		OK	00:11	

Data Download Date/Time: 01/16/2020 15:19

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASTINGS,KENNETH,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: HASTINGS,KENNETH
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 12/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: BILLINGS

DATE/TIME OF TEST: 12/27/2019 03:21:32

Diagnostics Check		OK	03:14
Air Blank		0.000	03:15
Control Test		0.081	03:15
Air Blank		0.000	03:15
Subject Sample	# 1	0.197	03:16
Breath Volume		1.640	03:16
Air Blank		0.000	03:16
Air Blank		0.000	03:18
Subject Sample	# 2	0.202	03:20
Breath Volume		2.886	03:20
Air Blank		0.000	03:20
Control Test		0.080	03:20
Air Blank		0.000	03:21
Diagnostics Check		OK	03:21

Data Download Date/Time: 01/16/2020 15:19

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 12/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: HUFF

DATE/TIME OF TEST: 12/27/2019 23:53:15

Diagnostics Check		OK	23:47
Air Blank		0.000	23:47
Control Test		0.081	23:47
Air Blank		0.000	23:48
Subject Sample	# 1	0.040	23:48
Breath Volume		1.792	23:48
Air Blank		0.000	23:49
Air Blank		0.000	23:51
Subject Sample	# 2	0.040	23:51
Breath Volume		1.847	23:51
Air Blank		0.000	23:52
Control Test		0.080	23:52
Air Blank		0.000	23:53
Diagnostics Check		OK	23:53

Data Download Date/Time: 01/16/2020 15:19

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASTINGS,KENNETH,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: HASTINGS,KENNETH
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 12/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 12/28/2019 04:15:01

Diagnostics Check		OK	04:08
Air Blank		0.000	04:08
Control Test		0.081	04:09
Air Blank		0.000	04:09
Subject Sample	# 1	0.127	04:10
Breath Volume		1.460	04:10
Air Blank		0.000	04:11
Air Blank		0.000	04:13
Subject Sample	# 2	0.127	04:13
Breath Volume		1.867	04:13
Air Blank		0.000	04:14
Control Test		0.080	04:14
Air Blank		0.000	04:14
Diagnostics Check		OK	04:15

Data Download Date/Time: 01/16/2020 15:19

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASTINGS,KENNETH,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: HASTINGS,KENNETH
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 12/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:44

SUBJECT'S LAST NAME: DEATHERAGE

DATE/TIME OF TEST: 12/31/2019 20:12:05

Diagnostics Check		OK	20:05
Air Blank		0.000	20:05
Control Test		0.080	20:06
Air Blank		0.000	20:06
Subject Sample	# 1	0.000	20:07
Breath Volume		1.988	20:07
Air Blank		0.000	20:07
Air Blank		0.000	20:09
Subject Sample	# 2	0.007	20:10
Breath Volume		2.292	20:10
Air Blank		0.000	20:11
Control Test		0.081	20:11
Air Blank		0.000	20:11
Diagnostics Check		OK	20:12

Data Download Date/Time: 01/16/2020 15:19

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEASE,MATTHEW,W
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: PEASE,MATTHEW
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 12/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: EZELL

DATE/TIME OF TEST: 12/29/2019 04:53:20

Diagnostics Check		OK	04:46
Air Blank		0.000	04:46
Control Test		0.080	04:46
Air Blank		0.000	04:47
Subject Sample	# 1	0.185	04:48
Breath Volume		3.960	04:48
Air Blank		0.000	04:49
Air Blank		0.000	04:50
Subject Sample	# 2	0.203	04:51
Breath Volume		4.777	04:51
Air Blank		0.000	04:52
Control Test		0.080	04:52
Air Blank		0.000	04:53
Diagnostics Check		OK	04:53

Data Download Date/Time: 01/15/2020 16:21

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARQUHARSON,OFC K,
BREATH TEST OPERATOR AGENCY: Plantation PD
ARRESTING OFFICER: FARQUHARSON,OFC K
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 11/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A5
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 05:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:25

SUBJECT'S LAST NAME: ROONOE

DATE/TIME OF TEST: 12/08/2019 05:53:00

Diagnostics Check		OK	05:46
Air Blank		0.000	05:46
Control Test		0.077	05:47
Air Blank		0.000	05:47
Subject Sample	# 1	0.174	05:48
Breath Volume		2.132	05:48
Air Blank		0.000	05:49
Air Blank		0.000	05:50
Subject Sample	# 2	0.172	05:51
Breath Volume		2.667	05:51
Air Blank		0.000	05:52
Control Test		0.076	05:52
Air Blank		0.000	05:52
Diagnostics Check		OK	05:53

Data Download Date/Time: 01/01/2020 13:53

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEAUREGARD,JEFF,
BREATH TEST OPERATOR AGENCY: Plantation PD
ARRESTING OFFICER: CORRERA,MADISON
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 12/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A5
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: OLIVO

DATE/TIME OF TEST: 12/28/2019 03:37:06

Diagnostics Check		OK	03:30
Air Blank		0.000	03:30
Control Test		0.077	03:31
Air Blank		0.000	03:31
Subject Sample	# 1	0.000	03:32
Breath Volume		2.972	03:32
Air Blank		0.000	03:32
Air Blank		0.000	03:34
Subject Sample	# 2	0.000	03:35
Breath Volume		3.265	03:35
Air Blank		0.000	03:36
Control Test		0.075	03:36
Air Blank		0.000	03:37
Diagnostics Check		OK	03:37

Data Download Date/Time: 01/01/2020 13:53

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DASHER,JOSHUA,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 12/02/2019 01:29:30

Diagnostics Check		OK	01:23
Air Blank		0.000	01:24
Control Test		0.080	01:24
Air Blank		0.000	01:24
Subject Sample	# 1	0.000	01:25
Breath Volume		1.667	01:25
Air Blank		0.000	01:25
Air Blank		0.000	01:27
Subject Sample	# 2	0.000	01:28
Breath Volume		1.644	01:28
Air Blank		0.000	01:28
Control Test		0.078	01:28
Air Blank		0.000	01:29
Diagnostics Check		OK	01:29

Data Download Date/Time: 12/03/2019 10:45

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: MELLON

DATE/TIME OF TEST: 12/04/2019 01:44:39

Diagnostics Check		OK	01:38
Air Blank		0.000	01:38
Control Test		0.079	01:39
Air Blank		0.000	01:39
Subject Sample	# 1	0.185	01:40
Breath Volume		1.187	01:40
Air Blank		0.000	01:40
Air Blank		0.000	01:42
Subject Sample	# 2	0.179	01:43
Breath Volume		1.312	01:43
Air Blank		0.000	01:43
Control Test		0.079	01:44
Air Blank		0.000	01:44
Diagnostics Check		OK	01:44

Data Download Date/Time: 12/20/2019 11:08

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FILBERT,E
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: ANDORFER

DATE/TIME OF TEST: 12/06/2019 02:58:03

Diagnostics Check		OK	02:51
Air Blank		0.000	02:52
Control Test		0.079	02:52
Air Blank		0.000	02:53
Subject Sample	# 1	0.136	02:53
Breath Volume		2.625	02:53
Air Blank		0.000	02:54
Air Blank		0.000	02:56
Subject Sample	# 2	0.124	02:56
Breath Volume		1.144	02:56
Air Blank		0.000	02:57
Control Test		0.079	02:57
Air Blank		0.000	02:57
Diagnostics Check		OK	02:58

Data Download Date/Time: 12/20/2019 11:08

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 12/06/2019 22:15:20

Diagnostics Check		OK	22:12	
Air Blank		0.000	22:12	
Control Test		0.079	22:13	
Air Blank		0.000	22:13	
Subject Sample	# 1		22:13	Subject Test Refused
Breath Volume		0.000	22:13	
Air Blank		0.000	22:14	
Control Test		0.080	22:14	
Air Blank		0.000	22:15	
Diagnostics Check		OK	22:15	

Data Download Date/Time: 12/20/2019 11:08

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: SCRANTON

DATE/TIME OF TEST: 12/06/2019 23:16:46

Diagnostics Check		OK	23:13	
Air Blank		0.000	23:14	
Control Test		0.079	23:14	
Air Blank		0.000	23:15	
Subject Sample	# 1		23:15	Subject Test Refused
Breath Volume		0.000	23:15	
Air Blank		0.000	23:15	
Control Test		0.079	23:16	
Air Blank		0.000	23:16	
Diagnostics Check		OK	23:16	

Data Download Date/Time: 12/20/2019 11:08

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: ALSTON

DATE/TIME OF TEST: 12/07/2019 02:04:32

Diagnostics Check		OK	01:58
Air Blank		0.000	01:59
Control Test		0.079	01:59
Air Blank		0.000	01:59
Subject Sample	# 1	0.000	02:00
Breath Volume		3.894	02:00
Air Blank		0.000	02:00
Air Blank		0.000	02:02
Subject Sample	# 2	0.000	02:03
Breath Volume		3.812	02:03
Air Blank		0.000	02:03
Control Test		0.080	02:03
Air Blank		0.000	02:04
Diagnostics Check		OK	02:04

Data Download Date/Time: 12/20/2019 11:09

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: ABT

DATE/TIME OF TEST: 12/07/2019 03:06:00

Diagnostics Check		OK	02:59
Air Blank		0.000	02:59
Control Test		0.079	03:00
Air Blank		0.000	03:00
Subject Sample	# 1	0.181	03:01
Breath Volume		2.460	03:01
Air Blank		0.000	03:02
Air Blank		0.000	03:03
Subject Sample	# 2	0.178	03:04
Breath Volume		1.214	03:04
Air Blank		0.000	03:05
Control Test		0.077	03:05
Air Blank		0.000	03:05
Diagnostics Check		OK	03:06

Data Download Date/Time: 12/20/2019 11:09

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: VILCHIS LOPEZ

DATE/TIME OF TEST: 12/07/2019 03:59:34

Diagnostics Check		OK	03:53
Air Blank		0.000	03:53
Control Test		0.079	03:54
Air Blank		0.000	03:54
Subject Sample	# 1	0.106	03:55
Breath Volume		3.996	03:55
Air Blank		0.000	03:55
Air Blank		0.000	03:57
Subject Sample	# 2	0.104	03:58
Breath Volume		3.730	03:58
Air Blank		0.000	03:58
Control Test		0.078	03:58
Air Blank		0.000	03:59
Diagnostics Check		OK	03:59

Data Download Date/Time: 12/20/2019 11:09

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TINGLE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: GRAY

DATE/TIME OF TEST: 12/07/2019 05:20:31

Diagnostics Check		OK	05:17	
Air Blank		0.000	05:18	
Control Test		0.079	05:18	
Air Blank		0.000	05:18	
Subject Sample	# 1		05:19	Subject Test Refused
Breath Volume		0.000	05:19	
Air Blank		0.000	05:19	
Control Test		0.080	05:19	
Air Blank		0.000	05:20	
Diagnostics Check		OK	05:20	

Data Download Date/Time: 12/20/2019 11:09

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: BADILLO-PAEZ

DATE/TIME OF TEST: 12/08/2019 02:16:37

Diagnostics Check		OK	02:13	
Air Blank		0.000	02:14	
Control Test		0.079	02:14	
Air Blank		0.000	02:15	
Subject Sample	# 1		02:15	Subject Test Refused
Breath Volume		0.000	02:15	
Air Blank		0.000	02:15	
Control Test		0.080	02:16	
Air Blank		0.000	02:16	
Diagnostics Check		OK	02:16	

Data Download Date/Time: 12/20/2019 11:09

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: BANKS

DATE/TIME OF TEST: 12/08/2019 05:22:17

Diagnostics Check		OK	05:16
Air Blank		0.000	05:16
Control Test		0.080	05:16
Air Blank		0.000	05:17
Subject Sample	# 1	0.108	05:17
Breath Volume		3.437	05:17
Air Blank		0.000	05:18
Air Blank		0.000	05:20
Subject Sample	# 2	0.109	05:20
Breath Volume		3.808	05:20
Air Blank		0.000	05:21
Control Test		0.079	05:21
Air Blank		0.000	05:22
Diagnostics Check		OK	05:22

Data Download Date/Time: 12/20/2019 11:09

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JENSEN,D
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: REED

DATE/TIME OF TEST: 12/08/2019 21:58:29

Diagnostics Check		OK	21:55	
Air Blank		0.000	21:56	
Control Test		0.080	21:56	
Air Blank		0.000	21:56	
Subject Sample	# 1		21:57	Subject Test Refused
Breath Volume		0.000	21:57	
Air Blank		0.000	21:57	
Control Test		0.079	21:57	
Air Blank		0.000	21:58	
Diagnostics Check		OK	21:58	

Data Download Date/Time: 12/20/2019 11:09

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: FLEMING

DATE/TIME OF TEST: 12/09/2019 02:11:23

Diagnostics Check		OK	02:05
Air Blank		0.000	02:05
Control Test		0.079	02:05
Air Blank		0.000	02:06
Subject Sample	# 1	0.106	02:06
Breath Volume		1.429	02:06
Air Blank		0.000	02:07
Air Blank		0.000	02:09
Subject Sample	# 2	0.110	02:09
Breath Volume		1.371	02:09
Air Blank		0.000	02:10
Control Test		0.079	02:10
Air Blank		0.000	02:11
Diagnostics Check		OK	02:11

Data Download Date/Time: 12/20/2019 11:09

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GIBSON,D
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: HALFORD

DATE/TIME OF TEST: 12/09/2019 03:25:33

Diagnostics Check		OK	03:19
Air Blank		0.000	03:19
Control Test		0.078	03:20
Air Blank		0.000	03:20
Subject Sample	# 1	0.151	03:21
Breath Volume		2.925	03:21
Air Blank		0.000	03:21
Air Blank		0.000	03:23
Subject Sample	# 2	0.150	03:24
Breath Volume		2.937	03:24
Air Blank		0.000	03:24
Control Test		0.078	03:24
Air Blank		0.000	03:25
Diagnostics Check		OK	03:25

Data Download Date/Time: 12/20/2019 11:09

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JORDAN,A
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: WEDDING

DATE/TIME OF TEST: 12/10/2019 22:55:08

Diagnostics Check		OK	22:48
Air Blank		0.000	22:49
Control Test		0.078	22:49
Air Blank		0.000	22:50
Subject Sample	# 1	0.115	22:50
Breath Volume		3.394	22:50
Air Blank		0.000	22:51
Air Blank		0.000	22:53
Subject Sample	# 2	0.110	22:53
Breath Volume		2.949	22:53
Air Blank		0.000	22:54
Control Test		0.079	22:54
Air Blank		0.000	22:55
Diagnostics Check		OK	22:55

Data Download Date/Time: 12/20/2019 11:09

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: RODGERS

DATE/TIME OF TEST: 12/10/2019 23:37:27

Diagnostics Check		OK	23:31
Air Blank		0.000	23:31
Control Test		0.078	23:31
Air Blank		0.000	23:32
Subject Sample	# 1	0.293	23:32
Breath Volume		2.195	23:32
Air Blank		0.000	23:33
Air Blank		0.000	23:35
Subject Sample	# 2	0.294	23:35
Breath Volume		2.992	23:35
Air Blank		0.000	23:36
Control Test		0.078	23:36
Air Blank		0.000	23:37
Diagnostics Check		OK	23:37

Data Download Date/Time: 12/20/2019 11:09

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 12/11/2019 04:09:32

Diagnostics Check		OK	04:06	
Air Blank		0.000	04:07	
Control Test		0.080	04:07	
Air Blank		0.000	04:08	
Subject Sample	# 1		04:08	Subject Test Refused
Breath Volume		0.000	04:08	
Air Blank		0.000	04:08	
Control Test		0.080	04:08	
Air Blank		0.000	04:09	
Diagnostics Check		OK	04:09	

Data Download Date/Time: 12/20/2019 11:10

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:29

SUBJECT'S LAST NAME: MAYS

DATE/TIME OF TEST: 12/11/2019 20:52:15

Diagnostics Check		OK	20:49	
Air Blank		0.000	20:49	
Control Test		0.079	20:50	
Air Blank		0.000	20:50	
Subject Sample	# 1		20:50	Subject Test Refused
Breath Volume		0.000	20:50	
Air Blank		0.000	20:51	
Control Test		0.079	20:51	
Air Blank		0.000	20:52	
Diagnostics Check		OK	20:52	

Data Download Date/Time: 12/20/2019 11:10

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: KIMBLE

DATE/TIME OF TEST: 12/12/2019 02:31:45

Diagnostics Check		OK	02:25
Air Blank		0.000	02:26
Control Test		0.080	02:26
Air Blank		0.000	02:26
Subject Sample	# 1	0.179	02:27
Breath Volume		4.203	02:27
Air Blank		0.000	02:27
Air Blank		0.000	02:29
Subject Sample	# 2	0.188	02:30
Breath Volume		4.089	02:30
Air Blank		0.000	02:30
Control Test		0.079	02:31
Air Blank		0.000	02:31
Diagnostics Check		OK	02:31

Data Download Date/Time: 12/20/2019 11:10

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: NAVARRO-REYES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: BASFORD

DATE/TIME OF TEST: 12/13/2019 03:16:26

Diagnostics Check		OK	03:10
Air Blank		0.000	03:10
Control Test		0.079	03:11
Air Blank		0.000	03:11
Subject Sample	# 1	0.164	03:12
Breath Volume		3.093	03:12
Air Blank		0.000	03:12
Air Blank		0.000	03:14
Subject Sample	# 2	0.162	03:14
Breath Volume		3.097	03:14
Air Blank		0.000	03:15
Control Test		0.079	03:15
Air Blank		0.000	03:16
Diagnostics Check		OK	03:16

Data Download Date/Time: 12/20/2019 11:10

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PRINCE,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: LONG

DATE/TIME OF TEST: 12/14/2019 02:16:46

Diagnostics Check		OK	02:13	
Air Blank		0.000	02:14	
Control Test		0.079	02:14	
Air Blank		0.000	02:15	
Subject Sample	# 1		02:15	Subject Test Refused
Breath Volume		0.000	02:15	
Air Blank		0.000	02:15	
Control Test		0.078	02:16	
Air Blank		0.000	02:16	
Diagnostics Check		OK	02:16	

Data Download Date/Time: 12/20/2019 11:10

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 12/15/2019 03:46:01

Diagnostics Check		OK	03:40
Air Blank		0.000	03:40
Control Test		0.079	03:40
Air Blank		0.000	03:41
Subject Sample	# 1	0.085	03:41
Breath Volume		1.878	03:41
Air Blank		0.000	03:42
Air Blank		0.000	03:44
Subject Sample	# 2	0.088	03:44
Breath Volume		2.074	03:44
Air Blank		0.000	03:45
Control Test		0.080	03:45
Air Blank		0.000	03:45
Diagnostics Check		OK	03:46

Data Download Date/Time: 12/20/2019 11:10

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: REEVES

DATE/TIME OF TEST: 12/16/2019 00:11:27

Diagnostics Check		OK	00:05
Air Blank		0.000	00:05
Control Test		0.079	00:06
Air Blank		0.000	00:06
Subject Sample	# 1	0.086	00:07
Breath Volume		2.699	00:07
Air Blank		0.000	00:07
Air Blank		0.000	00:09
Subject Sample	# 2	0.085	00:09
Breath Volume		2.871	00:09
Air Blank		0.000	00:10
Control Test		0.079	00:10
Air Blank		0.000	00:11
Diagnostics Check		OK	00:11

Data Download Date/Time: 12/20/2019 11:10

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DZAMKO,J
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: BENFIELD

DATE/TIME OF TEST: 12/17/2019 01:42:12

Diagnostics Check		OK	01:39	
Air Blank		0.000	01:39	
Control Test		0.079	01:40	
Air Blank		0.000	01:40	
Subject Sample	# 1		01:40	Subject Test Refused
Breath Volume		0.000	01:40	
Air Blank		0.000	01:41	
Control Test		0.079	01:41	
Air Blank		0.000	01:42	
Diagnostics Check		OK	01:42	

Data Download Date/Time: 12/20/2019 11:10

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JENSEN,D
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: PARRY

DATE/TIME OF TEST: 12/17/2019 04:46:16

Diagnostics Check		OK	04:43	
Air Blank		0.000	04:43	
Control Test		0.078	04:44	
Air Blank		0.000	04:44	
Subject Sample	# 1		04:44	Subject Test Refused
Breath Volume		0.000	04:44	
Air Blank		0.000	04:45	
Control Test		0.079	04:45	
Air Blank		0.000	04:46	
Diagnostics Check		OK	04:46	

Data Download Date/Time: 12/20/2019 11:10

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNS,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 12/18/2019 00:47:41

Diagnostics Check		OK	00:41
Air Blank		0.000	00:42
Control Test		0.079	00:42
Air Blank		0.000	00:43
Subject Sample	# 1	0.000	00:43
Breath Volume		1.390	00:43
Air Blank		0.000	00:43
Air Blank		0.000	00:45
Subject Sample	# 2	0.000	00:46
Breath Volume		1.777	00:46
Air Blank		0.000	00:46
Control Test		0.079	00:47
Air Blank		0.000	00:47
Diagnostics Check		OK	00:47

Data Download Date/Time: 12/20/2019 11:10

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: IKRAM,Y
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 12/18/2019 03:14:32

Diagnostics Check		OK	03:11	
Air Blank		0.000	03:12	
Control Test		0.079	03:12	
Air Blank		0.000	03:12	
Subject Sample	# 1		03:13	Subject Test Refused
Breath Volume		0.000	03:13	
Air Blank		0.000	03:13	
Control Test		0.080	03:13	
Air Blank		0.000	03:14	
Diagnostics Check		OK	03:14	

Data Download Date/Time: 12/20/2019 11:10

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: BOWMAN

DATE/TIME OF TEST: 12/20/2019 01:35:23

Diagnostics Check		OK	01:29
Air Blank		0.000	01:29
Control Test		0.080	01:29
Air Blank		0.000	01:30
Subject Sample	# 1	0.136	01:31
Breath Volume		1.148	01:31
Air Blank		0.000	01:31
Air Blank		0.000	01:33
Subject Sample	# 2	0.140	01:33
Breath Volume		1.269	01:33
Air Blank		0.000	01:34
Control Test		0.079	01:34
Air Blank		0.000	01:35
Diagnostics Check		OK	01:35

Data Download Date/Time: 12/20/2019 11:10

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: CONROY

DATE/TIME OF TEST: 12/21/2019 02:25:47

Diagnostics Check		OK	02:19
Air Blank		0.000	02:19
Control Test		0.080	02:19
Air Blank		0.000	02:20
Subject Sample	# 1	0.248	02:21
Breath Volume		1.761	02:21
Air Blank		0.000	02:21
Air Blank		0.000	02:23
Subject Sample	# 2	0.228	02:24
Breath Volume		1.628	02:24
Air Blank		0.000	02:24
Control Test		0.078	02:25
Air Blank		0.000	02:25
Diagnostics Check		OK	02:25

Data Download Date/Time: 01/16/2020 07:35

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNS,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 12/21/2019 03:58:04

Diagnostics Check		OK	03:51
Air Blank		0.000	03:52
Control Test		0.079	03:52
Air Blank		0.000	03:53
Subject Sample	# 1	0.108	03:53
Breath Volume		1.945	03:53
Air Blank		0.000	03:54
Air Blank		0.000	03:55
Subject Sample	# 2	0.109	03:56
Breath Volume		1.984	03:56
Air Blank		0.000	03:57
Control Test		0.079	03:57
Air Blank		0.000	03:57
Diagnostics Check		OK	03:58

Data Download Date/Time: 01/16/2020 07:35

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 12/21/2019 23:46:43

Diagnostics Check		OK	23:43	
Air Blank		0.000	23:44	
Control Test		0.080	23:44	
Air Blank		0.000	23:45	
Subject Sample	# 1		23:45	Subject Test Refused
Breath Volume		0.000	23:45	
Air Blank		0.000	23:45	
Control Test		0.079	23:46	
Air Blank		0.000	23:46	
Diagnostics Check		OK	23:46	

Data Download Date/Time: 01/16/2020 07:35

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: MC RAE

DATE/TIME OF TEST: 12/22/2019 04:16:59

Diagnostics Check		OK	04:08
Air Blank		0.000	04:09
Control Test		0.078	04:09
Air Blank		0.000	04:10
Subject Sample	# 1	0.159	04:12
Breath Volume		1.117	04:12
Air Blank		0.000	04:12
Air Blank		0.000	04:14
Subject Sample	# 2	0.154	04:15
Breath Volume		1.609	04:15
Air Blank		0.000	04:16
Control Test		0.078	04:16
Air Blank		0.000	04:16
Diagnostics Check		OK	04:16

Data Download Date/Time: 01/16/2020 07:35

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 12/23/2019 02:13:40

Diagnostics Check		OK	02:07
Air Blank		0.000	02:08
Control Test		0.079	02:08
Air Blank		0.000	02:08
Subject Sample	# 1	0.211	02:09
Breath Volume		2.335	02:09
Air Blank		0.000	02:09
Air Blank		0.000	02:11
Subject Sample	# 2	0.222	02:12
Breath Volume		1.554	02:12
Air Blank		0.000	02:12
Control Test		0.079	02:13
Air Blank		0.000	02:13
Diagnostics Check		OK	02:13

Data Download Date/Time: 01/16/2020 07:35

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 12/24/2019 04:17:47

Diagnostics Check		OK	04:11
Air Blank		0.000	04:11
Control Test		0.079	04:11
Air Blank		0.000	04:12
Subject Sample	# 1	0.221	04:13
Breath Volume		3.886	04:13
Air Blank		0.000	04:13
Air Blank		0.000	04:15
Subject Sample	# 2	0.217	04:16
Breath Volume		3.515	04:16
Air Blank		0.000	04:16
Control Test		0.079	04:17
Air Blank		0.000	04:17
Diagnostics Check		OK	04:17

Data Download Date/Time: 01/16/2020 07:35

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: DUNCAN

DATE/TIME OF TEST: 12/30/2019 00:21:15

Diagnostics Check		OK	00:18	
Air Blank		0.000	00:18	
Control Test		0.079	00:19	
Air Blank		0.000	00:19	
Subject Sample	# 1		00:19	Subject Test Refused
Breath Volume		0.000	00:19	
Air Blank		0.000	00:20	
Control Test		0.080	00:20	
Air Blank		0.000	00:21	
Diagnostics Check		OK	00:21	

Data Download Date/Time: 01/16/2020 07:35

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: YOUNG,S
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 12/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: DRESSELL

DATE/TIME OF TEST: 12/31/2019 01:33:58

Diagnostics Check		OK	01:31	
Air Blank		0.000	01:31	
Control Test		0.079	01:31	
Air Blank		0.000	01:32	
Subject Sample	# 1		01:32	Subject Test Refused
Breath Volume		0.000	01:32	
Air Blank		0.000	01:33	
Control Test		0.080	01:33	
Air Blank		0.000	01:33	
Diagnostics Check		OK	01:33	

Data Download Date/Time: 01/16/2020 07:35

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINCOLN,TREY,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LINCOLN,TREY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: KELLEY

DATE/TIME OF TEST: 12/12/2019 02:53:30

Diagnostics Check		OK	02:47
Air Blank		0.000	02:47
Control Test		0.080	02:47
Air Blank		0.000	02:48
Subject Sample	# 1	0.108	02:49
Breath Volume		2.054	02:49
Air Blank		0.000	02:49
Air Blank		0.000	02:51
Subject Sample	# 2	0.106	02:52
Breath Volume		2.312	02:52
Air Blank		0.000	02:52
Control Test		0.080	02:52
Air Blank		0.000	02:53
Diagnostics Check		OK	02:53

Data Download Date/Time: 01/02/2020 01:34

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,CALEB,D
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: RUSHWORTH,JOSEPH
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 12/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 22:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: BOWERS

DATE/TIME OF TEST: 12/30/2019 23:36:07

Diagnostics Check		OK	23:30
Air Blank		0.000	23:30
Control Test		0.079	23:30
Air Blank		0.000	23:31
Subject Sample	# 1	0.307	23:31
Breath Volume		1.273	23:31
Air Blank		0.000	23:32
Air Blank		0.000	23:34
Subject Sample	# 2	0.313	23:34
Breath Volume		1.312	23:34
Air Blank		0.000	23:35
Control Test		0.078	23:35
Air Blank		0.000	23:36
Diagnostics Check		OK	23:36

Data Download Date/Time: 01/30/2020 14:54

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLLINS,CAMERON,P
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: DEMING,WILLIAM
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 12/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: PEARCE

DATE/TIME OF TEST: 12/07/2019 02:14:19

Diagnostics Check		OK	02:07
Air Blank		0.000	02:07
Control Test		0.079	02:07
Air Blank		0.000	02:08
Subject Sample	# 1	0.212	02:09
Breath Volume		2.921	02:09
Air Blank		0.000	02:09
Air Blank		0.000	02:11
Subject Sample	# 2	0.213	02:12
Breath Volume		3.019	02:12
Air Blank		0.000	02:13
Control Test		0.079	02:13
Air Blank		0.000	02:14
Diagnostics Check		OK	02:14

Data Download Date/Time: 01/02/2020 11:17

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLER,DAVID,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: KELLER,DAVID
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 12/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:03

SUBJECT'S LAST NAME: HARRISON

DATE/TIME OF TEST: 12/10/2019 19:38:44

Diagnostics Check		OK	19:32
Air Blank		0.000	19:33
Control Test		0.079	19:33
Air Blank		0.000	19:33
Subject Sample	# 1	0.178	19:34
Breath Volume		1.429	19:34
Air Blank		0.000	19:34
Air Blank		0.000	19:36
Subject Sample	# 2	0.183	19:37
Breath Volume		1.265	19:37
Air Blank		0.000	19:37
Control Test		0.078	19:38
Air Blank		0.000	19:38
Diagnostics Check		OK	19:38

Data Download Date/Time: 01/02/2020 11:17

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FISCHER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: FISCHER,ANDREW
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 12/04/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:27

SUBJECT'S LAST NAME: CONTRERAS

DATE/TIME OF TEST: 12/14/2019 19:56:33

Diagnostics Check		OK	19:48
Air Blank		0.000	19:49
Control Test		0.079	19:49
Air Blank		0.000	19:50
Subject Sample	# 1	0.231	19:52
Breath Volume		1.609	19:52
Air Blank		0.000	19:52
Air Blank		0.000	19:54
Subject Sample	# 2	0.240	19:55
Breath Volume		1.222	19:55
Air Blank		0.000	19:55
Control Test		0.078	19:55
Air Blank		0.000	19:56
Diagnostics Check		OK	19:56

Data Download Date/Time: 01/02/2020 11:17

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAN,DONALD,E
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: DEAN,DONALD
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 12/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: RITTER

DATE/TIME OF TEST: 12/15/2019 05:06:03

Diagnostics Check		OK	05:00	
Air Blank		0.000	05:01	
Control Test		0.079	05:01	
Air Blank		0.000	05:01	
Subject Sample	# 1		05:04	Subject Test Refused
Breath Volume		0.050	05:04	
Air Blank		0.000	05:05	
Control Test		0.080	05:05	
Air Blank		0.000	05:05	
Diagnostics Check		OK	05:06	

Data Download Date/Time: 01/02/2020 11:17

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTHWAY,GEORGIA,K
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: NORTHWAY,GEORGIA
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 12/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 12/16/2019 05:33:00

Diagnostics Check		OK	05:26
Air Blank		0.000	05:26
Control Test		0.078	05:27
Air Blank		0.000	05:27
Subject Sample	# 1	0.055	05:28
Breath Volume		1.949	05:28
Air Blank		0.000	05:29
Air Blank		0.000	05:30
Subject Sample	# 2	0.052	05:31
Breath Volume		1.792	05:31
Air Blank		0.000	05:32
Control Test		0.077	05:32
Air Blank		0.000	05:32
Diagnostics Check		OK	05:33

Data Download Date/Time: 01/02/2020 11:17

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHULZE,JON,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: HEATH,SCOTT
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 12/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 12:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:10

SUBJECT'S LAST NAME: CHASTAIN

DATE/TIME OF TEST: 12/17/2019 13:43:36

Diagnostics Check		OK	13:37
Air Blank		0.000	13:37
Control Test		0.079	13:37
Air Blank		0.000	13:38
Subject Sample	# 1	0.000	13:39
Breath Volume		1.808	13:39
Air Blank		0.000	13:39
Air Blank		0.000	13:41
Subject Sample	# 2	0.000	13:42
Breath Volume		1.195	13:42
Air Blank		0.000	13:42
Control Test		0.079	13:42
Air Blank		0.000	13:43
Diagnostics Check		OK	13:43

Data Download Date/Time: 01/02/2020 11:17

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,GARY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 12/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:44

SUBJECT'S LAST NAME: ELLIARD

DATE/TIME OF TEST: 12/18/2019 19:16:46

Diagnostics Check		OK	19:10
Air Blank		0.000	19:11
Control Test		0.079	19:11
Air Blank		0.000	19:11
Subject Sample	# 1	0.030	19:12
Breath Volume		1.988	19:12
Air Blank		0.000	19:12
Air Blank		0.000	19:14
Subject Sample	# 2	0.029	19:15
Breath Volume		3.109	19:15
Air Blank		0.000	19:15
Control Test		0.080	19:16
Air Blank		0.000	19:16
Diagnostics Check		OK	19:16

Data Download Date/Time: 01/02/2020 11:17

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLER,DAVID,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: KELLER,DAVID
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 12/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: DEWBERRY

DATE/TIME OF TEST: 12/19/2019 02:20:25

Diagnostics Check		OK	02:13
Air Blank		0.000	02:14
Control Test		0.080	02:14
Air Blank		0.000	02:15
Subject Sample	# 1	0.147	02:15
Breath Volume		1.386	02:15
Air Blank		0.000	02:16
Air Blank		0.000	02:18
Subject Sample	# 2	0.156	02:18
Breath Volume		1.410	02:18
Air Blank		0.000	02:19
Control Test		0.080	02:19
Air Blank		0.000	02:20
Diagnostics Check		OK	02:20

Data Download Date/Time: 01/02/2020 11:17

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRADO,MARJORIE,D
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: PRADO,MARJORIE
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 12/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 15:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:50

SUBJECT'S LAST NAME: MCCLELLAN

DATE/TIME OF TEST: 12/28/2019 16:22:48

Diagnostics Check		OK	16:16
Air Blank		0.000	16:17
Control Test		0.078	16:17
Air Blank		0.000	16:17
Subject Sample	# 1	0.199	16:18
Breath Volume		3.609	16:18
Air Blank		0.000	16:19
Air Blank		0.000	16:20
Subject Sample	# 2	0.192	16:21
Breath Volume		3.066	16:21
Air Blank		0.000	16:21
Control Test		0.077	16:22
Air Blank		0.000	16:22
Diagnostics Check		OK	16:22

Data Download Date/Time: 01/02/2020 11:17

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENING,FRED,C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: TESLO,FRED
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 12/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 07:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:30

SUBJECT'S LAST NAME: GEORGE

DATE/TIME OF TEST: 12/31/2019 10:06:23

Diagnostics Check		OK	09:56
Air Blank		0.000	09:56
Control Test		0.080	09:56
Air Blank		0.000	09:57
Subject Sample	# 1	0.222	09:58
Breath Volume		2.375	09:58
Air Blank		0.000	09:59
Air Blank		0.000	10:01
Subject Sample	# 2	0.198	10:01
Breath Volume		2.468	10:01
Air Blank		0.000	10:02
Air Blank		0.000	10:04
Subject Sample	# 3	0.226	10:04
Breath Volume		2.367	10:04
Air Blank		0.000	10:05
Control Test		0.078	10:05
Air Blank		0.000	10:06
Diagnostics Check		OK	10:06

Data Download Date/Time: 01/02/2020 11:17

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001116

INSTRUMENT REGISTERED TO: EGLIN AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,CHRISTOPHER,P
BREATH TEST OPERATOR AGENCY: Eglin AFB PD
ARRESTING OFFICER: TRAMMELL,EZELL
ARRESTING OFFICER AGENCY: Eglin AFB PD
LAST AGENCY INSPECTION: 11/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: WICKER

DATE/TIME OF TEST: 12/14/2019 04:53:20

Diagnostics Check	OK	04:51	
Air Blank	0.000	04:52	
Control Test	0.079	04:52	
Air Blank		04:52	RFI Detect
Air Blank	0.000	04:53	

Data Download Date/Time: 12/18/2019 12:49

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001116

INSTRUMENT REGISTERED TO: EGLIN AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,CHRISTOPHER,P
BREATH TEST OPERATOR AGENCY: Eglin AFB PD
ARRESTING OFFICER: TRAMMELL,EZELL
ARRESTING OFFICER AGENCY: Eglin AFB PD
LAST AGENCY INSPECTION: 11/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: WICKER

DATE/TIME OF TEST: 12/14/2019 05:02:07

Diagnostics Check		OK	04:55
Air Blank		0.000	04:56
Control Test		0.079	04:56
Air Blank		0.000	04:56
Subject Sample	# 1	0.101	04:57
Breath Volume		4.140	04:57
Air Blank		0.000	04:58
Air Blank		0.000	05:00
Subject Sample	# 2	0.089	05:00
Breath Volume		2.152	05:00
Air Blank		0.000	05:01
Control Test		0.079	05:01
Air Blank		0.000	05:02
Diagnostics Check		OK	05:02

Data Download Date/Time: 12/18/2019 12:49

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRINIDAD,RICKY,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: TRINIDAD,RICKY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:02

SUBJECT'S LAST NAME: MARTINEZ SANZ

DATE/TIME OF TEST: 12/01/2019 06:39:12

Diagnostics Check		OK	06:32
Air Blank		0.000	06:33
Control Test		0.080	06:33
Air Blank		0.000	06:34
Subject Sample	# 1	0.186	06:34
Breath Volume		3.195	06:34
Air Blank		0.000	06:35
Air Blank		0.000	06:37
Subject Sample	# 2	0.191	06:37
Breath Volume		3.515	06:37
Air Blank		0.000	06:38
Control Test		0.080	06:38
Air Blank		0.000	06:39
Diagnostics Check		OK	06:39

Data Download Date/Time: 12/31/2019 18:20

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANNON,MATTHEW,F
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HANNON,MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 15:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:40

SUBJECT'S LAST NAME: WOLFF

DATE/TIME OF TEST: 12/03/2019 17:10:20

Diagnostics Check		OK	17:01	
Air Blank		0.000	17:02	
Control Test		0.080	17:02	
Air Blank		0.000	17:03	
Subject Sample	# 1		17:06	No Sample Provided
Breath Volume		0.000	17:06	
Air Blank		0.000	17:06	
Air Blank		0.000	17:08	
Subject Sample	# 2		17:08	Subject Test Refused
Breath Volume		0.000	17:08	
Air Blank		0.000	17:09	
Control Test		0.080	17:09	
Air Blank		0.000	17:10	
Diagnostics Check		OK	17:10	

Data Download Date/Time: 12/31/2019 18:20

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANKLIN,JOSEPH,B
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: FRANKLIN,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: GONZALEZ III

DATE/TIME OF TEST: 12/07/2019 04:07:51

Diagnostics Check		OK	04:01
Air Blank		0.000	04:01
Control Test		0.080	04:02
Air Blank		0.000	04:02
Subject Sample	# 1	0.202	04:03
Breath Volume		2.976	04:03
Air Blank		0.000	04:03
Air Blank		0.000	04:05
Subject Sample	# 2	0.199	04:06
Breath Volume		4.066	04:06
Air Blank		0.000	04:06
Control Test		0.079	04:07
Air Blank		0.000	04:07
Diagnostics Check		OK	04:07

Data Download Date/Time: 12/31/2019 18:20

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANKLIN,JOSEPH,B
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DIMARCO,AUSTIN
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 11/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: ARIAS-HERNANDEZ

DATE/TIME OF TEST: 12/07/2019 04:16:55

Diagnostics Check		OK	04:11	
Air Blank		0.000	04:12	
Control Test		0.079	04:12	
Air Blank		0.000	04:12	
Subject Sample	# 1		04:15	Subject Test Refused
Breath Volume		0.156	04:15	
Air Blank		0.000	04:16	
Control Test		0.079	04:16	
Air Blank		0.000	04:16	
Diagnostics Check		OK	04:16	

Data Download Date/Time: 12/31/2019 18:20

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANNON,MATTHEW,F
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HANNON,MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 14:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:53

SUBJECT'S LAST NAME: ZIMMERMAN

DATE/TIME OF TEST: 12/07/2019 18:20:35

Diagnostics Check		OK	18:14
Air Blank		0.000	18:14
Control Test		0.080	18:15
Air Blank		0.000	18:15
Subject Sample	# 1	0.000	18:16
Breath Volume		1.937	18:16
Air Blank		0.000	18:16
Air Blank		0.000	18:18
Subject Sample	# 2	0.000	18:19
Breath Volume		2.171	18:19
Air Blank		0.000	18:19
Control Test		0.080	18:19
Air Blank		0.000	18:20
Diagnostics Check		OK	18:20

Data Download Date/Time: 12/31/2019 18:20

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,BARRY,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: NIXON,BARRY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 11/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 08:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:32

SUBJECT'S LAST NAME: NUNEZ GARCIA

DATE/TIME OF TEST: 12/08/2019 10:04:57

Diagnostics Check		OK	09:53	
Air Blank		0.000	09:53	
Control Test		0.080	09:54	
Air Blank		0.000	09:54	
Subject Sample	# 1	0.097	09:57	Volume Not Met
Breath Volume		0.167	09:57	
Air Blank		0.000	09:58	
Air Blank		0.000	10:00	
Subject Sample	# 2	0.107	10:00	
Breath Volume		2.054	10:00	
Air Blank		0.000	10:01	
Air Blank		0.000	10:03	
Subject Sample	# 3	0.113	10:03	
Breath Volume		1.355	10:03	
Air Blank		0.000	10:04	
Control Test		0.079	10:04	
Air Blank		0.000	10:04	
Diagnostics Check		OK	10:04	

Data Download Date/Time: 12/31/2019 18:20

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHANSEN,LLOYD,D
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: HAMMELMAN,TYLER
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 11/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:08

SUBJECT'S LAST NAME: METZEL-COCIO

DATE/TIME OF TEST: 12/09/2019 05:34:39

Diagnostics Check		OK	05:28
Air Blank		0.000	05:28
Control Test		0.080	05:29
Air Blank		0.000	05:29
Subject Sample	# 1	0.056	05:30
Breath Volume		1.757	05:30
Air Blank		0.000	05:30
Air Blank		0.000	05:32
Subject Sample	# 2	0.050	05:33
Breath Volume		1.804	05:33
Air Blank		0.000	05:33
Control Test		0.078	05:34
Air Blank		0.000	05:34
Diagnostics Check		OK	05:34

Data Download Date/Time: 12/31/2019 18:20

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLINT,KILIAN,A
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: BLICKER,EVAN
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 11/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 06:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:28

SUBJECT'S LAST NAME: GARRISON

DATE/TIME OF TEST: 12/09/2019 07:55:38

Diagnostics Check		OK	07:49
Air Blank		0.000	07:49
Control Test		0.080	07:49
Air Blank		0.000	07:50
Subject Sample	# 1	0.210	07:51
Breath Volume		3.066	07:51
Air Blank		0.000	07:51
Air Blank		0.000	07:53
Subject Sample	# 2	0.196	07:54
Breath Volume		2.496	07:54
Air Blank		0.000	07:54
Control Test		0.079	07:55
Air Blank		0.000	07:55
Diagnostics Check		OK	07:55

Data Download Date/Time: 12/31/2019 18:20

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANKLIN,JOSEPH,B
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: FRANKLIN,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: GOBER

DATE/TIME OF TEST: 12/12/2019 03:30:20

Diagnostics Check		OK	03:24
Air Blank		0.000	03:24
Control Test		0.079	03:24
Air Blank		0.000	03:25
Subject Sample	# 1	0.186	03:25
Breath Volume		4.312	03:25
Air Blank		0.000	03:26
Air Blank		0.000	03:28
Subject Sample	# 2	0.184	03:28
Breath Volume		4.734	03:28
Air Blank		0.000	03:29
Control Test		0.080	03:29
Air Blank		0.000	03:30
Diagnostics Check		OK	03:30

Data Download Date/Time: 12/31/2019 18:20

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHANSEN,LLOYD,D
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: DIMARCO,AUSTIN
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 11/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 12/13/2019 00:35:11

Diagnostics Check		OK	00:27
Air Blank		0.000	00:28
Control Test		0.080	00:28
Air Blank		0.000	00:29
Subject Sample	# 1	0.063	00:30
Breath Volume		4.429	00:30
Air Blank		0.000	00:31
Air Blank		0.000	00:32
Subject Sample	# 2	0.062	00:33
Breath Volume		4.679	00:33
Air Blank		0.000	00:34
Control Test		0.078	00:34
Air Blank		0.000	00:35
Diagnostics Check		OK	00:35

Data Download Date/Time: 12/31/2019 18:20

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEQUEIRA,NATHAN,C
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SEQUEIRA,NATHAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:40

SUBJECT'S LAST NAME: HOVANCE

DATE/TIME OF TEST: 12/13/2019 17:09:34

Diagnostics Check		OK	17:00
Air Blank		0.000	17:01
Control Test		0.079	17:01
Air Blank		0.000	17:02
Subject Sample	# 1	0.235	17:05
Breath Volume		1.753	17:05
Air Blank		0.000	17:05
Air Blank		0.000	17:07
Subject Sample	# 2	0.231	17:08
Breath Volume		1.203	17:08
Air Blank		0.000	17:08
Control Test		0.077	17:08
Air Blank		0.000	17:09
Diagnostics Check		OK	17:09

Data Download Date/Time: 12/31/2019 18:20

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRINIDAD,RICKY,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: TRINIDAD,RICKY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: KLINE

DATE/TIME OF TEST: 12/14/2019 04:23:35

Diagnostics Check		OK	04:16
Air Blank		0.000	04:17
Control Test		0.079	04:17
Air Blank		0.000	04:18
Subject Sample	# 1	0.214	04:19
Breath Volume		1.667	04:19
Air Blank		0.000	04:19
Air Blank		0.000	04:21
Subject Sample	# 2	0.215	04:21
Breath Volume		2.003	04:21
Air Blank		0.000	04:22
Control Test		0.080	04:22
Air Blank		0.000	04:23
Diagnostics Check		OK	04:23

Data Download Date/Time: 12/31/2019 18:20

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUNNER,JOSEPH,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: RUNNER,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: KRINER

DATE/TIME OF TEST: 12/14/2019 05:07:54

Diagnostics Check		OK	05:01
Air Blank		0.000	05:01
Control Test		0.079	05:01
Air Blank		0.000	05:02
Subject Sample	# 1	0.138	05:03
Breath Volume		2.695	05:03
Air Blank		0.000	05:03
Air Blank		0.000	05:05
Subject Sample	# 2	0.145	05:06
Breath Volume		3.156	05:06
Air Blank		0.000	05:06
Control Test		0.079	05:07
Air Blank		0.000	05:07
Diagnostics Check		OK	05:07

Data Download Date/Time: 12/31/2019 18:20

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUNNER,JOSEPH,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: RUNNER,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: CONOLLY

DATE/TIME OF TEST: 12/14/2019 21:30:02

Diagnostics Check		OK	21:21
Air Blank		0.000	21:22
Control Test		0.080	21:22
Air Blank		0.000	21:22
Subject Sample	# 1	0.160	21:24
Breath Volume		1.593	21:24
Air Blank		0.000	21:25
Air Blank		0.000	21:27
Subject Sample	# 2	0.157	21:28
Breath Volume		1.800	21:28
Air Blank		0.000	21:29
Control Test		0.078	21:29
Air Blank		0.000	21:29
Diagnostics Check		OK	21:30

Data Download Date/Time: 12/31/2019 18:21

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,RAEFORD,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GRIFFIN,RAEFORD
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: RHODES

DATE/TIME OF TEST: 12/14/2019 22:04:20

Diagnostics Check		OK	21:56
Air Blank		0.000	21:56
Control Test		0.080	21:57
Air Blank		0.000	21:57
Subject Sample	# 1	0.133	21:59
Breath Volume		2.117	21:59
Air Blank		0.000	22:00
Air Blank		0.000	22:02
Subject Sample	# 2	0.131	22:02
Breath Volume		2.277	22:02
Air Blank		0.000	22:03
Control Test		0.079	22:03
Air Blank		0.000	22:04
Diagnostics Check		OK	22:04

Data Download Date/Time: 12/31/2019 18:21

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUNNER,JOSEPH,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: RUNNER,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/20/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: IGWE

DATE/TIME OF TEST: 12/15/2019 01:43:29

Diagnostics Check		OK	01:36
Air Blank		0.000	01:36
Control Test		0.079	01:36
Air Blank		0.000	01:37
Subject Sample	# 1	0.166	01:38
Breath Volume		2.726	01:38
Air Blank		0.000	01:39
Air Blank		0.000	01:41
Subject Sample	# 2	0.152	01:41
Breath Volume		1.867	01:41
Air Blank		0.000	01:42
Control Test		0.078	01:42
Air Blank		0.000	01:43
Diagnostics Check		OK	01:43

Data Download Date/Time: 12/31/2019 18:21

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,RAEFORD,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GRIFFIN,RAEFORD
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: ARGERENON

DATE/TIME OF TEST: 12/15/2019 03:56:43

Diagnostics Check		OK	03:46
Air Blank		0.000	03:46
Control Test		0.080	03:47
Air Blank		0.000	03:47
Subject Sample	# 1	0.141	03:48
Breath Volume		1.472	03:48
Air Blank		0.000	03:49
Air Blank		0.000	03:51
Subject Sample	# 2	0.164	03:51
Breath Volume		2.687	03:51
Air Blank		0.000	03:52
Air Blank		0.000	03:54
Subject Sample	# 3	0.149	03:55
Breath Volume		2.386	03:55
Air Blank		0.000	03:55
Control Test		0.080	03:56
Air Blank		0.000	03:56
Diagnostics Check		OK	03:56

Data Download Date/Time: 12/31/2019 18:21

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GABLE,DIANE,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: NOTO,JOHN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 14:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:47

SUBJECT'S LAST NAME: GABLE

DATE/TIME OF TEST: 12/15/2019 15:56:14

Diagnostics Check		OK	15:53	
Air Blank		0.000	15:54	
Control Test		0.080	15:54	
Air Blank		0.000	15:55	
Subject Sample	# 1		15:55	Sequence Aborted
Breath Volume		0.000	15:55	
Air Blank		0.000	15:56	

Data Download Date/Time: 12/31/2019 18:21

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GABLE,DIANE,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: NOTO,JOHN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/20/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 14:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:47

SUBJECT'S LAST NAME: GABLE

DATE/TIME OF TEST: 12/15/2019 16:16:23

Diagnostics Check		OK	16:07
Air Blank		0.000	16:08
Control Test		0.079	16:08
Air Blank		0.000	16:09
Subject Sample	# 1	0.093	16:11
Breath Volume		1.503	16:11
Air Blank		0.000	16:12
Air Blank		0.000	16:14
Subject Sample	# 2	0.098	16:14
Breath Volume		2.171	16:14
Air Blank		0.000	16:15
Control Test		0.079	16:15
Air Blank		0.000	16:16
Diagnostics Check		OK	16:16

Data Download Date/Time: 12/31/2019 18:21

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,TYLER,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KIRK,TYLER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: ROSARIO

DATE/TIME OF TEST: 12/16/2019 02:05:01

Diagnostics Check		OK	02:02	
Air Blank		0.000	02:02	
Control Test		0.079	02:02	
Air Blank		0.000	02:03	
Subject Sample	# 1		02:03	Subject Test Refused
Breath Volume		0.000	02:03	
Air Blank		0.000	02:04	
Control Test		0.079	02:04	
Air Blank		0.000	02:04	
Diagnostics Check		OK	02:05	

Data Download Date/Time: 12/31/2019 18:21

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,BARRY,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: NIXON,BARRY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 11/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 13:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:16

SUBJECT'S LAST NAME: CARRANZA

DATE/TIME OF TEST: 12/16/2019 14:42:55

Diagnostics Check		OK	14:36
Air Blank		0.000	14:37
Control Test		0.080	14:37
Air Blank		0.000	14:38
Subject Sample	# 1	0.058	14:38
Breath Volume		1.617	14:38
Air Blank		0.000	14:39
Air Blank		0.000	14:41
Subject Sample	# 2	0.056	14:41
Breath Volume		2.179	14:41
Air Blank		0.000	14:41
Control Test		0.079	14:42
Air Blank		0.000	14:42
Diagnostics Check		OK	14:42

Data Download Date/Time: 12/31/2019 18:21

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONNELL,JACOB,B
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: CONNELL,JACOB
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: POWELL

DATE/TIME OF TEST: 12/17/2019 03:57:15

Diagnostics Check		OK	03:50
Air Blank		0.000	03:51
Control Test		0.080	03:51
Air Blank		0.000	03:52
Subject Sample	# 1	0.200	03:52
Breath Volume		2.714	03:52
Air Blank		0.000	03:53
Air Blank		0.000	03:55
Subject Sample	# 2	0.193	03:55
Breath Volume		1.972	03:55
Air Blank		0.000	03:56
Control Test		0.080	03:56
Air Blank		0.000	03:57
Diagnostics Check		OK	03:57

Data Download Date/Time: 12/31/2019 18:21

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NETTLES,TRISTAN,O
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: NETTLES,TRISTAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: ROCK

DATE/TIME OF TEST: 12/18/2019 01:49:36

Diagnostics Check		OK	01:41	
Air Blank		0.000	01:41	
Control Test		0.079	01:41	
Air Blank		0.000	01:42	
Subject Sample	# 1		01:45	No Sample Provided
Breath Volume		0.000	01:45	
Air Blank		0.000	01:46	
Air Blank		0.000	01:48	
Subject Sample	# 2		01:48	Subject Test Refused
Breath Volume		0.000	01:48	
Air Blank		0.000	01:48	
Control Test		0.080	01:48	
Air Blank		0.000	01:49	
Diagnostics Check		OK	01:49	

Data Download Date/Time: 12/31/2019 18:21

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,RAEFORD,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GRIFFIN,RAEFORD
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: ALDRIDGE

DATE/TIME OF TEST: 12/19/2019 00:52:55

Diagnostics Check		OK	00:44	
Air Blank		0.000	00:44	
Control Test		0.080	00:44	
Air Blank		0.000	00:45	
Subject Sample	# 1		00:48	No Sample Provided
Breath Volume		0.000	00:48	
Air Blank		0.000	00:49	
Air Blank		0.000	00:51	
Subject Sample	# 2		00:51	Subject Test Refused
Breath Volume		0.000	00:51	
Air Blank		0.000	00:51	
Control Test		0.080	00:52	
Air Blank		0.000	00:52	
Diagnostics Check		OK	00:52	

Data Download Date/Time: 12/31/2019 18:21

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,RAEFORD,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GRIFFIN,RAEFORD
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: SANTANA

DATE/TIME OF TEST: 12/20/2019 00:35:11

Diagnostics Check		OK	00:28
Air Blank		0.000	00:29
Control Test		0.080	00:29
Air Blank		0.000	00:30
Subject Sample	# 1	0.183	00:30
Breath Volume		1.445	00:30
Air Blank		0.000	00:31
Air Blank		0.000	00:33
Subject Sample	# 2	0.180	00:33
Breath Volume		1.519	00:33
Air Blank		0.000	00:34
Control Test		0.080	00:34
Air Blank		0.000	00:35
Diagnostics Check		OK	00:35

Data Download Date/Time: 12/31/2019 18:21

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRINIDAD,RICKY,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: TRINIDAD,RICKY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: CARDENAS

DATE/TIME OF TEST: 12/20/2019 04:01:52

Diagnostics Check		OK	03:55
Air Blank		0.000	03:56
Control Test		0.080	03:56
Air Blank		0.000	03:56
Subject Sample	# 1	0.123	03:57
Breath Volume		3.218	03:57
Air Blank		0.000	03:58
Air Blank		0.000	03:59
Subject Sample	# 2	0.126	04:00
Breath Volume		3.152	04:00
Air Blank		0.000	04:00
Control Test		0.080	04:01
Air Blank		0.000	04:01
Diagnostics Check		OK	04:01

Data Download Date/Time: 12/31/2019 18:21

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANNON,MATTHEW,F
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HANNON,MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 14:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:03

SUBJECT'S LAST NAME: BALDRIDGE

DATE/TIME OF TEST: 12/20/2019 15:34:49

Diagnostics Check		OK	15:28
Air Blank		0.000	15:29
Control Test		0.080	15:29
Air Blank		0.000	15:30
Subject Sample	# 1	0.000	15:30
Breath Volume		2.449	15:30
Air Blank		0.000	15:31
Air Blank		0.000	15:33
Subject Sample	# 2	0.000	15:33
Breath Volume		2.355	15:33
Air Blank		0.000	15:33
Control Test		0.080	15:34
Air Blank		0.000	15:34
Diagnostics Check		OK	15:34

Data Download Date/Time: 12/31/2019 18:21

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONNELL,JACOB,B
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: CONNELL,JACOB
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:22

SUBJECT'S LAST NAME: SADORF

DATE/TIME OF TEST: 12/22/2019 05:53:52

Diagnostics Check		OK	05:45	
Air Blank		0.000	05:45	
Control Test		0.081	05:46	
Air Blank		0.000	05:46	
Subject Sample	# 1		05:49	No Sample Provided
Breath Volume		0.000	05:49	
Air Blank		0.000	05:50	
Air Blank		0.000	05:52	
Subject Sample	# 2		05:52	Subject Test Refused
Breath Volume		0.000	05:52	
Air Blank		0.000	05:52	
Control Test		0.081	05:53	
Air Blank		0.000	05:53	
Diagnostics Check		OK	05:53	

Data Download Date/Time: 12/31/2019 18:21

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,RAEFORD,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GRIFFIN,RAEFORD
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: HAOUATI

DATE/TIME OF TEST: 12/24/2019 05:38:56

Diagnostics Check		OK	05:31
Air Blank		0.000	05:32
Control Test		0.079	05:32
Air Blank		0.000	05:32
Subject Sample	# 1	0.167	05:34
Breath Volume		2.656	05:34
Air Blank		0.000	05:34
Air Blank		0.000	05:36
Subject Sample	# 2	0.169	05:37
Breath Volume		3.378	05:37
Air Blank		0.000	05:38
Control Test		0.078	05:38
Air Blank		0.000	05:38
Diagnostics Check		OK	05:38

Data Download Date/Time: 12/31/2019 18:21

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEQUEIRA,NATHAN,C
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SEQUEIRA,NATHAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:25

SUBJECT'S LAST NAME: COLON

DATE/TIME OF TEST: 12/24/2019 15:52:29

Diagnostics Check		OK	15:45
Air Blank		0.000	15:46
Control Test		0.079	15:46
Air Blank		0.000	15:47
Subject Sample	# 1	0.387	15:47
Breath Volume		4.003	15:47
Air Blank		0.000	15:48
Air Blank		0.000	15:50
Subject Sample	# 2	0.394	15:50
Breath Volume		3.402	15:50
Air Blank		0.000	15:51
Control Test		0.078	15:51
Air Blank		0.000	15:52
Diagnostics Check		OK	15:52

Data Download Date/Time: 12/31/2019 18:21

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEQUEIRA,NATHAN,C
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: CAMPANA,NICHOLAS
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: ROWE

DATE/TIME OF TEST: 12/24/2019 19:58:44

Diagnostics Check		OK	19:52
Air Blank		0.000	19:52
Control Test		0.079	19:53
Air Blank		0.000	19:53
Subject Sample	# 1	0.234	19:54
Breath Volume		3.628	19:54
Air Blank		0.000	19:54
Air Blank		0.000	19:56
Subject Sample	# 2	0.236	19:57
Breath Volume		4.000	19:57
Air Blank		0.000	19:57
Control Test		0.079	19:58
Air Blank		0.000	19:58
Diagnostics Check		OK	19:58

Data Download Date/Time: 12/31/2019 18:21

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NETTLES,TRISTAN,O
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: NETTLES,TRISTAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: HESTER

DATE/TIME OF TEST: 12/26/2019 04:21:15

Diagnostics Check		OK	04:12	
Air Blank		0.000	04:13	
Control Test		0.080	04:13	
Air Blank		0.000	04:14	
Subject Sample	# 1		04:17	No Sample Provided
Breath Volume		0.000	04:17	
Air Blank		0.000	04:17	
Air Blank		0.000	04:19	
Subject Sample	# 2		04:19	Subject Test Refused
Breath Volume		0.000	04:19	
Air Blank		0.000	04:20	
Control Test		0.081	04:20	
Air Blank		0.000	04:21	
Diagnostics Check		OK	04:21	

Data Download Date/Time: 12/31/2019 18:21

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANNON,MATTHEW,F
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HANNON,MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 08:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:30

SUBJECT'S LAST NAME: RECCHIO

DATE/TIME OF TEST: 12/26/2019 09:15:40

Diagnostics Check		OK	09:07	
Air Blank		0.000	09:07	
Control Test		0.079	09:07	
Air Blank		0.000	09:08	
Subject Sample	# 1		09:11	No Sample Provided
Breath Volume		0.000	09:11	
Air Blank		0.000	09:12	
Air Blank		0.000	09:14	
Subject Sample	# 2		09:14	Subject Test Refused
Breath Volume		0.000	09:14	
Air Blank		0.000	09:14	
Control Test		0.079	09:15	
Air Blank		0.000	09:15	
Diagnostics Check		OK	09:15	

Data Download Date/Time: 12/31/2019 18:21

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NETTLES,TRISTAN,O
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: NETTLES,TRISTAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: CEESAY

DATE/TIME OF TEST: 12/28/2019 03:22:47

Diagnostics Check		OK	03:16
Air Blank		0.000	03:16
Control Test		0.080	03:17
Air Blank		0.000	03:17
Subject Sample	# 1	0.126	03:18
Breath Volume		1.632	03:18
Air Blank		0.000	03:18
Air Blank		0.000	03:20
Subject Sample	# 2	0.128	03:21
Breath Volume		2.531	03:21
Air Blank		0.000	03:21
Control Test		0.079	03:22
Air Blank		0.000	03:22
Diagnostics Check		OK	03:22

Data Download Date/Time: 12/31/2019 18:21

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,BARRY,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: NIXON,BARRY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 12/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: CLOUD

DATE/TIME OF TEST: 12/28/2019 23:06:18

Diagnostics Check		OK	22:57
Air Blank		0.000	22:57
Control Test		0.079	22:57
Air Blank		0.000	22:58
Subject Sample	# 1	0.271	22:58
Breath Volume		1.500	22:58
Air Blank		0.000	22:59
Air Blank		0.000	23:01
Subject Sample	# 2	0.244	23:01
Breath Volume		1.636	23:01
Air Blank		0.000	23:02
Air Blank		0.000	23:04
Subject Sample	# 3	0.265	23:04
Breath Volume		2.335	23:04
Air Blank		0.000	23:05
Control Test		0.079	23:05
Air Blank		0.000	23:06
Diagnostics Check		OK	23:06

Data Download Date/Time: 12/31/2019 18:21

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NETTLES,TRISTAN,O
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: NETTLES,TRISTAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: GALLAHER

DATE/TIME OF TEST: 12/29/2019 05:35:53

Diagnostics Check		OK	05:28
Air Blank		0.000	05:29
Control Test		0.078	05:29
Air Blank		0.000	05:29
Subject Sample	# 1	0.132	05:30
Breath Volume		1.574	05:30
Air Blank		0.000	05:31
Air Blank		0.000	05:33
Subject Sample	# 2	0.149	05:34
Breath Volume		1.421	05:34
Air Blank		0.000	05:34
Control Test		0.078	05:35
Air Blank		0.000	05:35
Diagnostics Check		OK	05:35

Data Download Date/Time: 12/31/2019 18:21

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUNNER,JOSEPH,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: RUNNER,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: VANDRUFF

DATE/TIME OF TEST: 12/29/2019 21:57:38

Diagnostics Check		OK	21:46	
Air Blank		0.000	21:46	
Control Test		0.079	21:46	
Air Blank		0.000	21:47	
Subject Sample	# 1		21:50	No Sample Provided
Breath Volume		0.000	21:50	
Air Blank		0.000	21:51	
Air Blank		0.000	21:53	
Subject Sample	# 2		21:56	No Sample Provided
Breath Volume		0.000	21:56	
Air Blank		0.000	21:56	
Control Test		0.080	21:57	
Air Blank		0.000	21:57	
Diagnostics Check		OK	21:57	

Data Download Date/Time: 12/31/2019 18:21

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELLS,AUSTIN,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WELLS,AUSTIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: FERRARA

DATE/TIME OF TEST: 12/29/2019 22:31:14

Diagnostics Check		OK	22:24
Air Blank		0.000	22:25
Control Test		0.080	22:25
Air Blank		0.000	22:25
Subject Sample	# 1	0.185	22:26
Breath Volume		4.558	22:26
Air Blank		0.000	22:27
Air Blank		0.000	22:29
Subject Sample	# 2	0.188	22:29
Breath Volume		3.347	22:29
Air Blank		0.000	22:30
Control Test		0.080	22:30
Air Blank		0.000	22:31
Diagnostics Check		OK	22:31

Data Download Date/Time: 12/31/2019 18:21

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRINIDAD,RICKY,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: TRINIDAD,RICKY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: PEREZGOMEZ

DATE/TIME OF TEST: 12/31/2019 00:36:11

Diagnostics Check		OK	00:29
Air Blank		0.000	00:29
Control Test		0.079	00:30
Air Blank		0.000	00:30
Subject Sample	# 1	0.139	00:31
Breath Volume		2.468	00:31
Air Blank		0.000	00:32
Air Blank		0.000	00:34
Subject Sample	# 2	0.135	00:34
Breath Volume		1.597	00:34
Air Blank		0.000	00:35
Control Test		0.080	00:35
Air Blank		0.000	00:36
Diagnostics Check		OK	00:36

Data Download Date/Time: 12/31/2019 18:21

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMPSON,TREY,A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: GAYDOS,SAVANNAH
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 16:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:01

SUBJECT'S LAST NAME: COOKE

DATE/TIME OF TEST: 12/29/2019 17:30:22

Diagnostics Check	OK	17:29	
Air Blank	0.000	17:29	
Control Test	0.000	17:29	Control Outside Tolerance
Air Blank	0.000	17:30	

Data Download Date/Time: 01/02/2020 02:36

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMPSON,TREY,A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: GAYDOS,SAVANNAH
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 16:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:01

SUBJECT'S LAST NAME: COOKE

DATE/TIME OF TEST: 12/29/2019 17:33:30

Diagnostics Check	OK	17:32	
Air Blank	0.000	17:32	
Control Test	0.000	17:32	Control Outside Tolerance
Air Blank	0.000	17:33	

Data Download Date/Time: 01/02/2020 02:36

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMPSON,TREY,A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: GAYDOS,SAVANNAH
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 16:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:01

SUBJECT'S LAST NAME: COOKE

DATE/TIME OF TEST: 12/29/2019 17:47:22

Diagnostics Check	OK	17:46	
Air Blank	0.000	17:46	
Control Test	0.000	17:46	Control Outside Tolerance
Air Blank	0.000	17:47	

Data Download Date/Time: 01/02/2020 02:36

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMPSON,TREY,A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: GAYDOS,SAVANNAH
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 16:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:01

SUBJECT'S LAST NAME: COOKE

DATE/TIME OF TEST: 12/29/2019 17:50:24

Diagnostics Check	OK	17:49	
Air Blank	0.000	17:49	
Control Test	0.000	17:49	Control Outside Tolerance
Air Blank	0.000	17:50	

Data Download Date/Time: 01/02/2020 02:36

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMPSON,TREY,A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: GAYDOS,SAVANNAH
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 16:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:01

SUBJECT'S LAST NAME: COOKE

DATE/TIME OF TEST: 12/29/2019 17:58:55

Diagnostics Check	OK	17:57	
Air Blank	0.000	17:58	
Control Test	0.000	17:58	Control Outside Tolerance
Air Blank	0.000	17:58	

Data Download Date/Time: 01/02/2020 02:36

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMPSON,TREY,A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: GAYDOS,SAVANNAH
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 16:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:01

SUBJECT'S LAST NAME: COOKE

DATE/TIME OF TEST: 12/29/2019 18:08:56

Diagnostics Check	OK	18:07	
Air Blank	0.000	18:08	
Control Test	0.000	18:08	Control Outside Tolerance
Air Blank	0.000	18:08	

Data Download Date/Time: 01/02/2020 02:36

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMPSON,TREY,A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: GAYDOS,SAVANNAH
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 16:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:01

SUBJECT'S LAST NAME: COOKE

DATE/TIME OF TEST: 12/29/2019 18:15:14

Diagnostics Check	OK	18:13	
Air Blank	0.000	18:14	
Control Test	0.000	18:14	Control Outside Tolerance
Air Blank	0.000	18:15	

Data Download Date/Time: 01/02/2020 02:36

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMPSON,TREY,A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: GAYDOS,SAVANNAH
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 16:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:01

SUBJECT'S LAST NAME: COOKE

DATE/TIME OF TEST: 12/29/2019 18:23:20

Diagnostics Check	OK	18:22	
Air Blank	0.000	18:22	
Control Test	0.000	18:22	Control Outside Tolerance
Air Blank	0.000	18:23	

Data Download Date/Time: 01/02/2020 02:36

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMPSON,TREY,A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: GAYDOS,SAVANNAH
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 16:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:01

SUBJECT'S LAST NAME: COOKE

DATE/TIME OF TEST: 12/29/2019 18:28:42

Diagnostics Check	OK	18:27	
Air Blank	0.000	18:27	
Control Test	0.000	18:28	Control Outside Tolerance
Air Blank	0.000	18:28	

Data Download Date/Time: 01/02/2020 02:36

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMPSON,TREY,A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: GAYDOS,SAVANNAH
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 16:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:01

SUBJECT'S LAST NAME: COOKE

DATE/TIME OF TEST: 12/29/2019 18:57:59

Diagnostics Check		OK	18:42	
Air Blank		0.000	18:43	
Control Test		0.081	18:43	
Air Blank		0.000	18:44	
Subject Sample	# 1	0.151	18:45	
Breath Volume		1.687	18:45	
Air Blank		0.000	18:45	
Air Blank		0.000	18:47	
Subject Sample	# 2	0.142	18:50	Slope Not Level
Breath Volume		1.152	18:50	
Air Blank		0.000	18:51	
Air Blank		0.000	18:53	
Subject Sample	# 3	0.127	18:56	Volume Not Met
Breath Volume		0.625	18:56	
Air Blank		0.000	18:57	
Control Test		0.079	18:57	
Air Blank		0.000	18:57	
Diagnostics Check		OK	18:57	

Data Download Date/Time: 01/02/2020 02:36

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMPSON,TREY,A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: GAYDOS,SAVANNAH
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 16:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:01

SUBJECT'S LAST NAME: COOKE

DATE/TIME OF TEST: 12/29/2019 19:21:32

Diagnostics Check		OK	19:14
Air Blank		0.000	19:15
Control Test		0.081	19:15
Air Blank		0.000	19:16
Subject Sample	# 1	0.152	19:16
Breath Volume		2.914	19:16
Air Blank		0.000	19:17
Air Blank		0.000	19:19
Subject Sample	# 2	0.144	19:19
Breath Volume		2.480	19:19
Air Blank		0.000	19:20
Control Test		0.081	19:20
Air Blank		0.000	19:21
Diagnostics Check		OK	19:21

Data Download Date/Time: 01/02/2020 02:36

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAYDOS,JONATHAN,A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: YOST,CRAIG
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 20:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:28

SUBJECT'S LAST NAME: ROXBY

DATE/TIME OF TEST: 12/29/2019 21:51:41

Diagnostics Check		OK	21:48	
Air Blank		0.000	21:49	
Control Test		0.080	21:49	
Air Blank		0.000	21:50	
Subject Sample	# 1		21:50	Subject Test Refused
Breath Volume		0.000	21:50	
Air Blank		0.000	21:50	
Control Test		0.081	21:51	
Air Blank		0.000	21:51	
Diagnostics Check		OK	21:51	

Data Download Date/Time: 01/02/2020 02:36

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAYDOS,JONATHAN,A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: YOST,CRAIG
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 20:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: ROXBY

DATE/TIME OF TEST: 12/29/2019 22:20:29

Diagnostics Check		OK	22:13
Air Blank		0.000	22:13
Control Test		0.081	22:13
Air Blank		0.000	22:14
Subject Sample	# 1	0.226	22:15
Breath Volume		2.214	22:15
Air Blank		0.000	22:16
Air Blank		0.000	22:18
Subject Sample	# 2	0.228	22:18
Breath Volume		2.324	22:18
Air Blank		0.000	22:19
Control Test		0.080	22:19
Air Blank		0.000	22:20
Diagnostics Check		OK	22:20

Data Download Date/Time: 01/02/2020 02:36

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001122

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JUDAH,DAVID,B
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: JUDAH,DAVID
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 11/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 12/08/2019 04:32:47

Diagnostics Check		OK	04:24
Air Blank		0.000	04:24
Control Test		0.079	04:25
Air Blank		0.000	04:25
Subject Sample	# 1	0.271	04:26
Breath Volume		1.500	04:26
Air Blank		0.000	04:27
Air Blank		0.000	04:29
Subject Sample	# 2	0.254	04:31
Breath Volume		2.074	04:31
Air Blank		0.000	04:31
Control Test		0.081	04:32
Air Blank		0.000	04:32
Diagnostics Check		OK	04:32

Data Download Date/Time: 12/30/2019 11:56

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001122

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARIAS,ELOY,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: GOLVERG,JOSHUA
ARRESTING OFFICER AGENCY: MICCOSUKEE PD
LAST AGENCY INSPECTION: 11/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 04:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: CAMEJO ESNARD

DATE/TIME OF TEST: 12/08/2019 05:49:40

Diagnostics Check		OK	05:38	
Air Blank		0.000	05:38	
Control Test		0.078	05:39	
Air Blank		0.000	05:39	
Subject Sample	# 1		05:42	No Sample Provided
Breath Volume		0.000	05:42	
Air Blank		0.000	05:43	
Air Blank		0.000	05:45	
Subject Sample	# 2		05:48	No Sample Provided
Breath Volume		0.000	05:48	
Air Blank		0.000	05:48	
Control Test		0.078	05:49	
Air Blank		0.000	05:49	
Diagnostics Check		OK	05:49	

Data Download Date/Time: 12/30/2019 11:56

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001122

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARIAS,ELOY,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: ARIAS,ELOY
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 11/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: SALINAS

DATE/TIME OF TEST: 12/09/2019 02:45:40

Diagnostics Check		OK	02:34	
Air Blank		0.000	02:34	
Control Test		0.079	02:35	
Air Blank		0.000	02:35	
Subject Sample	# 1	0.139	02:38	Volume Not Met
Breath Volume		0.425	02:38	
Air Blank		0.000	02:39	
Air Blank		0.000	02:41	
Subject Sample	# 2	0.150	02:44	
Breath Volume		1.160	02:44	
Air Blank			02:45	Purge Fail
Air Blank		0.000	02:45	

Data Download Date/Time: 12/30/2019 11:56

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001122

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANCHEZ,JUSTIN,A
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: SANCHEZ,JUSTIN
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 11/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: ECHEZARRETA

DATE/TIME OF TEST: 12/11/2019 22:23:27

Diagnostics Check		OK	22:11	
Air Blank		0.000	22:12	
Control Test		0.080	22:12	
Air Blank		0.000	22:13	
Subject Sample	# 1	0.154	22:16	Volume Not Met
Breath Volume		0.019	22:16	
Air Blank		0.000	22:17	
Air Blank		0.000	22:18	
Subject Sample	# 2	0.134	22:21	Volume Not Met
Breath Volume		0.089	22:21	
Air Blank		0.000	22:22	
Control Test		0.080	22:22	
Air Blank		0.000	22:23	
Diagnostics Check		OK	22:23	

Data Download Date/Time: 12/30/2019 11:57

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001122

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARIAS,ELOY,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: ARIAS,ELOY
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 11/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 12/13/2019 02:39:26

Diagnostics Check		OK	02:32
Air Blank		0.000	02:33
Control Test		0.080	02:33
Air Blank		0.000	02:34
Subject Sample	# 1	0.074	02:34
Breath Volume		2.984	02:34
Air Blank		0.000	02:35
Air Blank		0.000	02:37
Subject Sample	# 2	0.070	02:37
Breath Volume		3.480	02:37
Air Blank		0.000	02:38
Control Test		0.079	02:38
Air Blank		0.000	02:39
Diagnostics Check		OK	02:39

Data Download Date/Time: 12/30/2019 11:57

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001122

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACHADO,ERICK,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: MACHADO,ERICK
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 11/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 12/14/2019 01:57:39

Diagnostics Check		OK	01:46	
Air Blank		0.000	01:46	
Control Test		0.080	01:47	
Air Blank		0.000	01:47	
Subject Sample	# 1		01:50	No Sample Provided
Breath Volume		0.000	01:50	
Air Blank		0.000	01:51	
Air Blank		0.000	01:53	
Subject Sample	# 2		01:56	No Sample Provided
Breath Volume		0.000	01:56	
Air Blank		0.000	01:56	
Control Test		0.079	01:57	
Air Blank		0.000	01:57	
Diagnostics Check		OK	01:57	

Data Download Date/Time: 12/30/2019 11:57

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001122

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAXIMILIEN,RONALD,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: MAXIMILIEN,RONALD
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 11/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: MIRANDA

DATE/TIME OF TEST: 12/14/2019 02:49:42

Diagnostics Check		OK	02:46	
Air Blank		0.000	02:47	
Control Test		0.080	02:47	
Air Blank		0.000	02:47	
Subject Sample	# 1		02:48	Subject Test Refused
Breath Volume		0.000	02:48	
Air Blank		0.000	02:48	
Control Test		0.079	02:49	
Air Blank		0.000	02:49	
Diagnostics Check		OK	02:49	

Data Download Date/Time: 12/30/2019 11:57

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001122

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROMERO,OMAR,E
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: FABIEN,JEFFREY
ARRESTING OFFICER AGENCY: MIAMI SPRINGS PD
LAST AGENCY INSPECTION: 11/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 14:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:40

SUBJECT'S LAST NAME: JOSEPH

DATE/TIME OF TEST: 12/15/2019 16:20:42

Diagnostics Check		OK	16:08	
Air Blank		0.000	16:08	
Control Test		0.077	16:09	
Air Blank		0.000	16:09	
Subject Sample	# 1	0.240	16:10	
Breath Volume		1.777	16:10	
Air Blank		0.000	16:11	
Air Blank		0.000	16:12	
Subject Sample	# 2	0.245	16:16	Volume Not Met
Breath Volume		0.015	16:16	
Air Blank		0.000	16:17	
Air Blank		0.000	16:18	
Subject Sample	# 3	0.243	16:19	
Breath Volume		1.812	16:19	
Air Blank		0.000	16:19	
Control Test		0.077	16:20	
Air Blank		0.000	16:20	
Diagnostics Check		OK	16:20	

Data Download Date/Time: 12/30/2019 11:57

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001122

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAXIMILIEN,RONALD,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: MAXIMILIEN,RONALD
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 11/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: SOUZA

DATE/TIME OF TEST: 12/16/2019 04:29:55

Diagnostics Check		OK	04:21	
Air Blank		0.000	04:21	
Control Test		0.079	04:22	
Air Blank		0.000	04:22	
Subject Sample	# 1		04:25	No Sample Provided
Breath Volume		0.000	04:25	
Air Blank		0.000	04:26	
Air Blank		0.000	04:28	
Subject Sample	# 2		04:28	Subject Test Refused
Breath Volume		0.000	04:28	
Air Blank		0.000	04:29	
Control Test		0.079	04:29	
Air Blank		0.000	04:29	
Diagnostics Check		OK	04:29	

Data Download Date/Time: 12/30/2019 11:57

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001122

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOMINGUEZ,JOSE,F
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: DOMINGUEZ,JOSE
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 11/30/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 11:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: ABREU

DATE/TIME OF TEST: 12/20/2019 04:31:18

Diagnostics Check		OK	04:21	
Air Blank		0.000	04:22	
Control Test		0.081	04:22	
Air Blank		0.000	04:23	
Subject Sample	# 1		04:26	No Sample Provided
Breath Volume		0.000	04:26	
Air Blank		0.000	04:26	
Air Blank		0.000	04:28	
Subject Sample	# 2		04:29	Subject Test Refused
Breath Volume		0.000	04:29	
Air Blank		0.000	04:30	
Control Test		0.080	04:30	
Air Blank		0.000	04:31	
Diagnostics Check		OK	04:31	

Data Download Date/Time: 12/30/2019 11:57

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001122

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLO,DENNIS,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: GALLO,DENNIS
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 11/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 18:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:31

SUBJECT'S LAST NAME: RAMZAN

DATE/TIME OF TEST: 12/20/2019 22:34:04

Diagnostics Check		OK	22:27
Air Blank		0.000	22:27
Control Test		0.079	22:28
Air Blank		0.000	22:28
Subject Sample	# 1	0.236	22:29
Breath Volume		2.417	22:29
Air Blank		0.000	22:30
Air Blank		0.000	22:31
Subject Sample	# 2	0.241	22:32
Breath Volume		2.546	22:32
Air Blank		0.000	22:33
Control Test		0.079	22:33
Air Blank		0.000	22:33
Diagnostics Check		OK	22:34

Data Download Date/Time: 12/30/2019 11:57

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001122

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JUDAH,DAVID,B
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: JUDAH,DAVID
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 11/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 04:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: VASQUEZ

DATE/TIME OF TEST: 12/22/2019 05:24:03

Diagnostics Check		OK	05:17
Air Blank		0.000	05:17
Control Test		0.080	05:18
Air Blank		0.000	05:18
Subject Sample	# 1	0.166	05:19
Breath Volume		1.617	05:19
Air Blank		0.000	05:20
Air Blank		0.000	05:22
Subject Sample	# 2	0.166	05:22
Breath Volume		1.828	05:22
Air Blank		0.000	05:23
Control Test		0.079	05:23
Air Blank		0.000	05:23
Diagnostics Check		OK	05:24

Data Download Date/Time: 12/30/2019 11:57

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001122

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMIREZ,JOSHUA,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: RAMIREZ,JOSHUA
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 11/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: OWENS

DATE/TIME OF TEST: 12/26/2019 01:27:45

Diagnostics Check		OK	01:21
Air Blank		0.000	01:21
Control Test		0.080	01:22
Air Blank		0.000	01:22
Subject Sample	# 1	0.177	01:23
Breath Volume		1.410	01:23
Air Blank		0.000	01:24
Air Blank		0.000	01:25
Subject Sample	# 2	0.180	01:26
Breath Volume		1.261	01:26
Air Blank		0.000	01:26
Control Test		0.078	01:27
Air Blank		0.000	01:27
Diagnostics Check		OK	01:27

Data Download Date/Time: 12/30/2019 11:57

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001122

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOMINGUEZ,JOSE,F
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: DOMINGUEZ,JOSE
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 11/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: POZO

DATE/TIME OF TEST: 12/27/2019 01:40:54

Diagnostics Check		OK	01:34
Air Blank		0.000	01:34
Control Test		0.079	01:35
Air Blank		0.000	01:35
Subject Sample	# 1	0.157	01:36
Breath Volume		1.421	01:36
Air Blank		0.000	01:37
Air Blank		0.000	01:38
Subject Sample	# 2	0.154	01:39
Breath Volume		1.164	01:39
Air Blank		0.000	01:39
Control Test		0.078	01:40
Air Blank		0.000	01:40
Diagnostics Check		OK	01:40

Data Download Date/Time: 12/30/2019 11:57

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001123

INSTRUMENT REGISTERED TO: FHP TROOP D ORLANDO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,VICTOR,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: RIVERA,VICTOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:44

SUBJECT'S LAST NAME: COWARD

DATE/TIME OF TEST: 12/25/2019 21:11:36

Diagnostics Check		OK	21:05	
Air Blank		0.000	21:05	
Control Test		0.080	21:06	
Air Blank		0.000	21:06	
Subject Sample	# 1	0.123	21:07	
Breath Volume		2.328	21:07	
Air Blank		0.000	21:07	
Air Blank		0.000	21:09	
Subject Sample	# 2	0.119	21:10	
Breath Volume		1.218	21:10	
Air Blank		0.000	21:10	
Control Test		0.081	21:10	
Air Blank			21:11	RFI Detect
Air Blank		0.000	21:11	

Data Download Date/Time: 01/27/2020 23:42

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001123

INSTRUMENT REGISTERED TO: FHP TROOP D ORLANDO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRACH,STEFAN,G
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: DRACH,STEFAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: LAGOS MEJIA

DATE/TIME OF TEST: 12/28/2019 01:43:58

Diagnostics Check		OK	01:37	
Air Blank		0.000	01:38	
Control Test		0.080	01:38	
Air Blank		0.000	01:39	
Subject Sample	# 1	0.184	01:40	
Breath Volume		2.062	01:40	
Air Blank		0.000	01:40	
Air Blank		0.000	01:42	
Subject Sample	# 2	0.180	01:43	
Breath Volume		2.050	01:43	
Air Blank			01:43	RFI Detect
Air Blank		0.000	01:43	

Data Download Date/Time: 01/27/2020 23:42

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001123

INSTRUMENT REGISTERED TO: FHP TROOP D ORLANDO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRACH,STEFAN,G
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: DRACH,STEFAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: DOMINGUEZ

DATE/TIME OF TEST: 12/31/2019 22:03:17

Diagnostics Check		OK	21:57
Air Blank		0.000	21:57
Control Test		0.081	21:58
Air Blank		0.000	21:58
Subject Sample	# 1	0.097	21:58
Breath Volume		1.777	21:58
Air Blank		0.000	21:59
Air Blank		0.000	22:01
Subject Sample	# 2	0.095	22:01
Breath Volume		1.914	22:01
Air Blank		0.000	22:02
Control Test		0.081	22:02
Air Blank		0.000	22:03
Diagnostics Check		OK	22:03

Data Download Date/Time: 01/27/2020 23:42

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELCHER,ALEX,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: PROPST,JOHN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:48

SUBJECT'S LAST NAME: KEELEY

DATE/TIME OF TEST: 12/01/2019 20:22:18

Diagnostics Check		OK	20:10	
Air Blank		0.000	20:11	
Control Test		0.083	20:11	
Air Blank		0.000	20:12	
Subject Sample	# 1		20:15	No Sample Provided
Breath Volume		0.000	20:15	
Air Blank		0.000	20:15	
Air Blank		0.000	20:17	
Subject Sample	# 2		20:20	No Sample Provided
Breath Volume		0.000	20:20	
Air Blank		0.000	20:21	
Control Test		0.080	20:21	
Air Blank		0.000	20:22	
Diagnostics Check		OK	20:22	

Data Download Date/Time: 12/26/2019 13:14

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,ROBERT,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HAMILTON,ROBERT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: KESTING

DATE/TIME OF TEST: 12/03/2019 01:12:26

Diagnostics Check		OK	01:05	
Air Blank		0.000	01:06	
Control Test		0.083	01:06	
Air Blank		0.000	01:06	
Subject Sample	# 1	0.000	01:07	
Breath Volume		1.695	01:07	
Air Blank		0.000	01:08	
Air Blank		0.000	01:10	
Subject Sample	# 2	0.000	01:10	
Breath Volume		2.804	01:10	
Air Blank		0.000	01:11	
Control Test		0.084	01:11	
Air Blank		0.000	01:11	
Diagnostics Check		Fail	01:11	Analytical Stability Test
Air Blank		0.000	01:12	

Data Download Date/Time: 12/26/2019 13:14

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES,STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ALLRED,COBY
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: LOWE

DATE/TIME OF TEST: 12/03/2019 23:39:55

Diagnostics Check		OK	23:33
Air Blank		0.000	23:34
Control Test		0.083	23:34
Air Blank		0.000	23:35
Subject Sample	# 1	0.103	23:35
Breath Volume		1.679	23:35
Air Blank		0.000	23:36
Air Blank		0.000	23:37
Subject Sample	# 2	0.103	23:38
Breath Volume		1.218	23:38
Air Blank		0.000	23:39
Control Test		0.080	23:39
Air Blank		0.000	23:39
Diagnostics Check		OK	23:39

Data Download Date/Time: 12/26/2019 13:14

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELCHER,ALEX,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: BEDFORD

DATE/TIME OF TEST: 12/05/2019 01:02:02

Diagnostics Check		OK	00:55
Air Blank		0.000	00:55
Control Test		0.082	00:55
Air Blank		0.000	00:56
Subject Sample	# 1	0.183	00:57
Breath Volume		1.574	00:57
Air Blank		0.000	00:57
Air Blank		0.000	00:59
Subject Sample	# 2	0.198	01:00
Breath Volume		1.394	01:00
Air Blank		0.000	01:01
Control Test		0.082	01:01
Air Blank		0.000	01:01
Diagnostics Check		OK	01:02

Data Download Date/Time: 12/26/2019 13:14

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELCHER,ALEX,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: PROPST,JOHN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 12/05/2019 01:13:02

Diagnostics Check		OK	01:05
Air Blank		0.000	01:05
Control Test		0.081	01:06
Air Blank		0.000	01:06
Subject Sample	# 1	0.127	01:08
Breath Volume		1.667	01:08
Air Blank		0.000	01:09
Air Blank		0.000	01:11
Subject Sample	# 2	0.131	01:11
Breath Volume		1.566	01:11
Air Blank		0.000	01:12
Control Test		0.082	01:12
Air Blank		0.000	01:12
Diagnostics Check		OK	01:13

Data Download Date/Time: 12/26/2019 13:14

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES,STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SYKES,STEVEN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: FANORD

DATE/TIME OF TEST: 12/13/2019 03:18:39

Diagnostics Check		OK	03:09
Air Blank		0.000	03:09
Control Test		0.082	03:10
Air Blank		0.000	03:10
Subject Sample	# 1	0.146	03:12
Breath Volume		1.511	03:12
Air Blank		0.000	03:13
Air Blank		0.000	03:15
Subject Sample	# 2	0.146	03:17
Breath Volume		1.449	03:17
Air Blank		0.000	03:17
Control Test		0.082	03:18
Air Blank		0.000	03:18
Diagnostics Check		OK	03:18

Data Download Date/Time: 12/26/2019 13:14

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIGGINBOTHAM,JACK,C
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HIGGINBOTHAM,JACK
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: MYERS

DATE/TIME OF TEST: 12/13/2019 04:33:22

Diagnostics Check		OK	04:25
Air Blank		0.000	04:25
Control Test		0.082	04:25
Air Blank		0.000	04:26
Subject Sample	# 1	0.138	04:28
Breath Volume		1.257	04:28
Air Blank		0.000	04:29
Air Blank		0.000	04:31
Subject Sample	# 2	0.145	04:31
Breath Volume		2.906	04:31
Air Blank		0.000	04:32
Control Test		0.081	04:32
Air Blank		0.000	04:33
Diagnostics Check		OK	04:33

Data Download Date/Time: 12/26/2019 13:15

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELCHER,ALEX,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:17

SUBJECT'S LAST NAME: TALAVERA-CABAN

DATE/TIME OF TEST: 12/14/2019 20:58:00

Diagnostics Check		OK	20:44	
Air Blank		0.000	20:44	
Control Test		0.082	20:44	
Air Blank		0.000	20:45	
Subject Sample	# 1		20:48	No Sample Provided
Breath Volume		0.000	20:48	
Air Blank		0.000	20:49	
Air Blank		0.000	20:51	
Subject Sample	# 2	0.183	20:53	
Breath Volume		1.386	20:53	
Air Blank		0.000	20:53	
Air Blank		0.000	20:55	
Subject Sample	# 3	0.199	20:56	
Breath Volume		1.308	20:56	
Air Blank		0.000	20:57	
Control Test		0.079	20:57	
Air Blank		0.000	20:57	
Diagnostics Check		OK	20:58	

Data Download Date/Time: 12/26/2019 13:15

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIGGEMBOTHAM,JACK,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HIGGEMBOTHAM,JACK
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: MCGARY

DATE/TIME OF TEST: 12/16/2019 21:08:38

Diagnostics Check		OK	21:02
Air Blank		0.000	21:02
Control Test		0.081	21:02
Air Blank		0.000	21:03
Subject Sample	# 1	0.153	21:04
Breath Volume		1.289	21:04
Air Blank		0.000	21:04
Air Blank		0.000	21:06
Subject Sample	# 2	0.155	21:07
Breath Volume		1.816	21:07
Air Blank		0.000	21:07
Control Test		0.082	21:08
Air Blank		0.000	21:08
Diagnostics Check		OK	21:08

Data Download Date/Time: 12/26/2019 13:15

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELCHER,ALEX,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:26

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 12/19/2019 19:53:24

Diagnostics Check		OK	19:46
Air Blank		0.000	19:46
Control Test		0.082	19:47
Air Blank		0.000	19:47
Subject Sample	# 1	0.000	19:48
Breath Volume		2.714	19:48
Air Blank		0.000	19:49
Air Blank		0.000	19:51
Subject Sample	# 2	0.000	19:52
Breath Volume		2.667	19:52
Air Blank		0.000	19:52
Control Test		0.081	19:52
Air Blank		0.000	19:53
Diagnostics Check		OK	19:53

Data Download Date/Time: 12/26/2019 13:15

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELCHER,ALEX,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LONDON,JOHN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: CANNON

DATE/TIME OF TEST: 12/20/2019 01:42:41

Diagnostics Check		OK	01:31	
Air Blank		0.000	01:31	
Control Test		0.081	01:32	
Air Blank		0.000	01:32	
Subject Sample	# 1		01:35	No Sample Provided
Breath Volume		0.000	01:35	
Air Blank		0.000	01:36	
Air Blank		0.000	01:38	
Subject Sample	# 2		01:41	No Sample Provided
Breath Volume		0.000	01:41	
Air Blank		0.000	01:41	
Control Test		0.081	01:42	
Air Blank		0.000	01:42	
Diagnostics Check		OK	01:42	

Data Download Date/Time: 12/26/2019 13:15

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIGGINBOTHAM,JACK,C
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HIGGINBOTHAM,JACK
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:17

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 12/20/2019 17:59:10

Diagnostics Check		OK	17:47	
Air Blank		0.000	17:48	
Control Test		0.081	17:48	
Air Blank		0.000	17:49	
Subject Sample	# 1		17:52	No Sample Provided
Breath Volume		0.000	17:52	
Air Blank		0.000	17:52	
Air Blank		0.000	17:54	
Subject Sample	# 2		17:57	No Sample Provided
Breath Volume		0.000	17:57	
Air Blank		0.000	17:58	
Control Test		0.082	17:58	
Air Blank		0.000	17:59	
Diagnostics Check		OK	17:59	

Data Download Date/Time: 12/26/2019 13:15

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,DALE,E
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ROBERTS,DALE
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:11

SUBJECT'S LAST NAME: BRAVO

DATE/TIME OF TEST: 12/23/2019 22:37:34

Diagnostics Check		OK	22:34	
Air Blank		0.000	22:35	
Control Test		0.082	22:35	
Air Blank		0.000	22:35	
Subject Sample	# 1		22:36	Subject Test Refused
Breath Volume		0.000	22:36	
Air Blank		0.000	22:36	
Control Test		0.081	22:36	
Air Blank		0.000	22:37	
Diagnostics Check		OK	22:37	

Data Download Date/Time: 12/26/2019 13:15

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELCHER,ALEX,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: PROPST,JOHN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: RUIZ-REYES

DATE/TIME OF TEST: 12/28/2019 04:40:21

Diagnostics Check		OK	04:33
Air Blank		0.000	04:34
Control Test		0.081	04:34
Air Blank		0.000	04:34
Subject Sample	# 1	0.162	04:35
Breath Volume		1.273	04:35
Air Blank		0.000	04:36
Air Blank		0.000	04:38
Subject Sample	# 2	0.171	04:38
Breath Volume		1.773	04:38
Air Blank		0.000	04:39
Control Test		0.081	04:39
Air Blank		0.000	04:40
Diagnostics Check		OK	04:40

Data Download Date/Time: 03/03/2020 11:18

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JOHNSTON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 15:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:43

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 12/28/2019 17:16:58

Diagnostics Check		OK	17:07	
Air Blank		0.000	17:07	
Control Test		0.083	17:07	
Air Blank		0.000	17:08	
Subject Sample	# 1	0.151	17:11	Volume Not Met
Breath Volume		0.000	17:11	
Air Blank		0.000	17:12	
Air Blank		0.000	17:13	
Subject Sample	# 2		17:15	Subject Test Refused
Breath Volume		0.898	17:15	
Air Blank		0.000	17:16	
Control Test		0.081	17:16	
Air Blank		0.000	17:16	
Diagnostics Check		OK	17:16	

Data Download Date/Time: 03/03/2020 11:18

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,DALE,E
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ROBERTS,DALE
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: LUCAS

DATE/TIME OF TEST: 12/28/2019 19:10:43

Diagnostics Check		OK	19:01
Air Blank		0.000	19:02
Control Test		0.079	19:02
Air Blank		0.000	19:02
Subject Sample	# 1	0.265	19:05
Breath Volume		1.121	19:05
Air Blank		0.000	19:05
Air Blank		0.000	19:07
Subject Sample	# 2	0.269	19:09
Breath Volume		1.199	19:09
Air Blank		0.000	19:09
Control Test		0.081	19:10
Air Blank		0.000	19:10
Diagnostics Check		OK	19:10

Data Download Date/Time: 03/03/2020 11:18

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELCHER,ALEX,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: PROPST,JOHN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 12/29/2019 02:59:26

Diagnostics Check		OK	02:48	
Air Blank		0.000	02:48	
Control Test		0.081	02:48	
Air Blank		0.000	02:49	
Subject Sample	# 1		02:52	No Sample Provided
Breath Volume		0.000	02:52	
Air Blank		0.000	02:52	
Air Blank		0.000	02:54	
Subject Sample	# 2		02:58	No Sample Provided
Breath Volume		0.000	02:58	
Air Blank		0.000	02:58	
Control Test		0.083	02:58	
Air Blank		0.000	02:59	
Diagnostics Check		OK	02:59	

Data Download Date/Time: 03/03/2020 11:18

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001129

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,JARED,R
BREATH TEST OPERATOR AGENCY: Putnam County SO
ARRESTING OFFICER: DAVIS,JARED
ARRESTING OFFICER AGENCY: Putnam County SO
LAST AGENCY INSPECTION: 12/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A3
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 09:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:31

SUBJECT'S LAST NAME: HOOVER

DATE/TIME OF TEST: 12/07/2019 10:58:48

Diagnostics Check		OK	10:52
Air Blank		0.000	10:52
Control Test		0.081	10:53
Air Blank		0.000	10:53
Subject Sample	# 1	0.000	10:54
Breath Volume		2.195	10:54
Air Blank		0.000	10:54
Air Blank		0.000	10:56
Subject Sample	# 2	0.000	10:57
Breath Volume		2.125	10:57
Air Blank		0.000	10:57
Control Test		0.079	10:58
Air Blank		0.000	10:58
Diagnostics Check		OK	10:58

Data Download Date/Time: 01/02/2020 18:01

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001129

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GELSEY,RICHARD,D
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: GELSEY,RICHARD
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: MEDINA

DATE/TIME OF TEST: 12/15/2019 05:47:07

Diagnostics Check		OK	05:40
Air Blank		0.000	05:41
Control Test		0.081	05:41
Air Blank		0.000	05:42
Subject Sample	# 1	0.159	05:42
Breath Volume		3.406	05:42
Air Blank		0.000	05:43
Air Blank		0.000	05:45
Subject Sample	# 2	0.168	05:45
Breath Volume		2.730	05:45
Air Blank		0.000	05:46
Control Test		0.081	05:46
Air Blank		0.000	05:47
Diagnostics Check		OK	05:47

Data Download Date/Time: 01/02/2020 18:01

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRYAN,STEVEN,A
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: WALLIS,DEAN
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 11/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1048227
DRY GAS EXPIRY DATE: 11/15/2021
VIOLATION CODE: DUI
ARREST TIME: 04:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: KEITH

DATE/TIME OF TEST: 12/15/2019 05:19:30

Diagnostics Check		OK	05:12
Air Blank		0.000	05:12
Control Test		0.081	05:13
Air Blank		0.000	05:13
Subject Sample	# 1	0.150	05:14
Breath Volume		2.800	05:14
Air Blank		0.000	05:15
Air Blank		0.000	05:17
Subject Sample	# 2	0.152	05:18
Breath Volume		2.597	05:18
Air Blank		0.000	05:18
Control Test		0.081	05:18
Air Blank		0.000	05:19
Diagnostics Check		OK	05:19

Data Download Date/Time: 12/30/2019 14:33

Printed: 19 March 2020

COBRA ©1997-2014 CMI, Inc.