

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: NETTLES,SHARON
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 01/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 17:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: BEEBE

DATE/TIME OF TEST: 02/01/2019 23:11:40

Diagnostics Check	OK	23:05
Air Blank	0.000	23:06
Control Test	0.078	23:06
Air Blank	0.000	23:07
Subject Sample # 1	0.000	23:07
Breath Volume	1.367	23:07
Air Blank	0.000	23:07
Air Blank	0.000	23:09
Subject Sample # 2	0.000	23:10
Breath Volume	1.695	23:10
Air Blank	0.000	23:10
Control Test	0.077	23:11
Air Blank	0.000	23:11
Diagnostics Check	OK	23:11

Data Download Date/Time: 02/17/2019 20:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: IRIZARRY,CHRIS
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 01/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: RAMOS GABRIEL

DATE/TIME OF TEST: 02/01/2019 23:34:29

Diagnostics Check	OK	23:26
Air Blank	0.000	23:26
Control Test	0.077	23:27
Air Blank	0.000	23:27
Subject Sample # 1	0.268	23:29
Breath Volume	2.609	23:29
Air Blank	0.000	23:29
Air Blank	0.000	23:31
Subject Sample # 2	0.266	23:32
Breath Volume	2.226	23:32
Air Blank	0.000	23:33
Control Test	0.079	23:33
Air Blank	0.000	23:34
Diagnostics Check	OK	23:34

Data Download Date/Time: 02/17/2019 20:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: HOLIDAY,LUCAS
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 01/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: NUNEZ

DATE/TIME OF TEST: 02/02/2019 02:58:33

Diagnostics Check	OK	02:49
Air Blank	0.000	02:50
Control Test	0.078	02:50
Air Blank	0.000	02:51
Subject Sample # 1	0.120	02:53
Breath Volume	1.187	02:53
Air Blank	0.000	02:54
Air Blank	0.000	02:56
Subject Sample # 2	0.119	02:57
Breath Volume	1.187	02:57
Air Blank	0.000	02:57
Control Test	0.077	02:57
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

Data Download Date/Time: 02/17/2019 20:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: ANDREWS,GARY
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 01/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: RAMOS RODRIGUEZ

DATE/TIME OF TEST: 02/02/2019 03:34:22

Diagnostics Check	OK	03:28
Air Blank	0.000	03:28
Control Test	0.076	03:29
Air Blank	0.000	03:29
Subject Sample # 1	0.107	03:30
Breath Volume	4.449	03:30
Air Blank	0.000	03:30
Air Blank	0.000	03:32
Subject Sample # 2	0.113	03:32
Breath Volume	4.222	03:32
Air Blank	0.000	03:33
Control Test	0.077	03:33
Air Blank	0.000	03:34
Diagnostics Check	OK	03:34

Data Download Date/Time: 02/17/2019 20:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: ROSS,JACOB
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 01/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: FRYER

DATE/TIME OF TEST: 02/02/2019 03:48:31

Diagnostics Check	OK	03:40
Air Blank	0.000	03:41
Control Test	0.077	03:41
Air Blank	0.000	03:42
Subject Sample # 1	0.172	03:43
Breath Volume	1.839	03:43
Air Blank	0.000	03:44
Air Blank	0.000	03:46
Subject Sample # 2	0.172	03:47
Breath Volume	1.625	03:47
Air Blank	0.000	03:47
Control Test	0.077	03:47
Air Blank	0.000	03:48
Diagnostics Check	OK	03:48

Data Download Date/Time: 02/17/2019 20:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINKOPF,TIM,E
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: KINKOPF,TIM
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 01/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 05:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: JUAREZ MALAVE

DATE/TIME OF TEST: 02/06/2019 06:02:59

Diagnostics Check	OK	05:56
Air Blank	0.000	05:56
Control Test	0.077	05:56
Air Blank	0.000	05:57
Subject Sample # 1	0.169	05:58
Breath Volume	2.554	05:58
Air Blank	0.000	05:58
Air Blank	0.000	06:00
Subject Sample # 2	0.175	06:01
Breath Volume	3.261	06:01
Air Blank	0.000	06:02
Control Test	0.078	06:02
Air Blank	0.000	06:02
Diagnostics Check	OK	06:02

Data Download Date/Time: 02/17/2019 20:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: KEE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 01/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 02/08/2019 02:17:17

Diagnostics Check	OK	02:05	
Air Blank	0.000	02:06	
Control Test	0.078	02:06	
Air Blank	0.000	02:07	
Subject Sample	# 1 0.161	02:10	Volume Not Met
Breath Volume	0.015	02:10	
Air Blank	0.000	02:10	
Air Blank	0.000	02:12	
Subject Sample	# 2 0.170	02:15	Volume Not Met
Breath Volume	0.035	02:15	
Air Blank	0.000	02:16	
Control Test	0.080	02:16	
Air Blank	0.000	02:17	
Diagnostics Check	OK	02:17	

Data Download Date/Time: 02/17/2019 20:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: KEE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 01/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 02/08/2019 02:27:11

Diagnostics Check	OK	02:19
Air Blank	0.000	02:19
Control Test	0.078	02:19
Air Blank	0.000	02:20
Subject Sample # 1	0.170	02:21
Breath Volume	1.894	02:21
Air Blank	0.000	02:21
Air Blank	0.000	02:23
Subject Sample # 2	0.161	02:25
Breath Volume	1.246	02:25
Air Blank	0.000	02:26
Control Test	0.078	02:26
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

Data Download Date/Time: 02/17/2019 20:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: CANNADAY,ZACHARY
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 01/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: HOLMES

DATE/TIME OF TEST: 02/08/2019 03:51:34

Diagnostics Check	OK	03:39	
Air Blank	0.000	03:40	
Control Test	0.078	03:40	
Air Blank	0.000	03:40	
Subject Sample	# 1 0.125	03:43	Volume Not Met
Breath Volume	1.113	03:43	
Air Blank	0.000	03:44	
Air Blank	0.000	03:46	
Subject Sample	# 2 0.127	03:46	
Breath Volume	2.578	03:46	
Air Blank	0.000	03:47	
Air Blank	0.000	03:49	
Subject Sample	# 3 0.123	03:50	
Breath Volume	2.628	03:50	
Air Blank	0.000	03:50	
Control Test	0.080	03:50	
Air Blank	0.000	03:51	
Diagnostics Check	OK	03:51	

Data Download Date/Time: 02/17/2019 20:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: DEL CASTILLO,BRET
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 01/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 04:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 02/08/2019 05:48:35

Diagnostics Check	OK	05:35	
Air Blank	0.000	05:35	
Control Test	0.078	05:36	
Air Blank	0.000	05:36	
Subject Sample	# 1 0.144	05:39	Volume Not Met
Breath Volume	0.000	05:39	
Air Blank	0.000	05:40	
Air Blank	0.000	05:42	
Subject Sample	# 2 0.127	05:43	
Breath Volume	1.101	05:43	
Air Blank	0.000	05:44	
Air Blank	0.000	05:45	
Subject Sample	# 3 0.132	05:47	
Breath Volume	1.582	05:47	
Air Blank	0.000	05:47	
Control Test	0.079	05:48	
Air Blank	0.000	05:48	
Diagnostics Check	OK	05:48	

Data Download Date/Time: 02/17/2019 20:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIEBELL,DON,R
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: COCCHOILA,AUSTIN
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 02/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: MARCOTT

DATE/TIME OF TEST: 02/23/2019 01:18:15

Diagnostics Check	OK	01:07
Air Blank	0.000	01:07
Control Test	0.080	01:08
Air Blank	0.000	01:08
Subject Sample # 1	0.229	01:10
Breath Volume	1.980	01:10
Air Blank	0.000	01:11
Air Blank	0.000	01:13
Subject Sample # 2	0.226	01:16
Breath Volume	2.238	01:16
Air Blank	0.000	01:17
Control Test	0.078	01:17
Air Blank	0.000	01:18
Diagnostics Check	OK	01:18

Data Download Date/Time: 04/01/2019 07:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIEBELL,DON,R
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: TRIEBELL,DON
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 02/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: SEEDS

DATE/TIME OF TEST: 02/23/2019 04:42:45

Diagnostics Check	OK	04:35
Air Blank	0.000	04:36
Control Test	0.081	04:36
Air Blank	0.000	04:36
Subject Sample # 1	0.268	04:37
Breath Volume	1.222	04:37
Air Blank	0.000	04:38
Air Blank	0.000	04:40
Subject Sample # 2	0.266	04:41
Breath Volume	2.191	04:41
Air Blank	0.000	04:41
Control Test	0.080	04:42
Air Blank	0.000	04:42
Diagnostics Check	OK	04:42

Data Download Date/Time: 04/01/2019 07:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,ERIC,G
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: TRIEBELL,DONALD
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 02/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:11

SUBJECT'S LAST NAME: FERGUSON

DATE/TIME OF TEST: 02/23/2019 22:44:37

Diagnostics Check	OK	22:38
Air Blank	0.000	22:38
Control Test	0.080	22:38
Air Blank	0.000	22:39
Subject Sample # 1	0.185	22:39
Breath Volume	3.273	22:39
Air Blank	0.000	22:40
Air Blank	0.000	22:42
Subject Sample # 2	0.178	22:43
Breath Volume	3.191	22:43
Air Blank	0.000	22:43
Control Test	0.079	22:44
Air Blank	0.000	22:44
Diagnostics Check	OK	22:44

Data Download Date/Time: 04/01/2019 07:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIEBELL,DON,R
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: TRIEBELL,DON
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 02/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: ARMA

DATE/TIME OF TEST: 02/24/2019 02:31:27

Diagnostics Check	OK	02:24
Air Blank	0.000	02:25
Control Test	0.081	02:25
Air Blank	0.000	02:25
Subject Sample # 1	0.136	02:26
Breath Volume	2.074	02:26
Air Blank	0.000	02:27
Air Blank	0.000	02:28
Subject Sample # 2	0.129	02:29
Breath Volume	2.078	02:29
Air Blank	0.000	02:30
Control Test	0.079	02:30
Air Blank	0.000	02:31
Diagnostics Check	OK	02:31

Data Download Date/Time: 04/01/2019 07:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,ERIC,G
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: HINES,MICHAEL
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 02/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 15:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:48

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 02/24/2019 17:19:24

Diagnostics Check	OK	17:09
Air Blank	0.000	17:10
Control Test	0.080	17:10
Air Blank	0.000	17:11
Subject Sample # 1	0.286	17:14
Breath Volume	1.199	17:14
Air Blank	0.000	17:14
Air Blank	0.000	17:16
Subject Sample # 2	0.295	17:17
Breath Volume	1.183	17:17
Air Blank	0.000	17:18
Control Test	0.080	17:18
Air Blank	0.000	17:19
Diagnostics Check	OK	17:19

Data Download Date/Time: 04/01/2019 07:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCRANN,SEAN,M
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: PHILBIN,SKYLER
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 02/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: OWENS

DATE/TIME OF TEST: 02/26/2019 01:58:40

Diagnostics Check	OK	01:52
Air Blank	0.000	01:52
Control Test	0.081	01:53
Air Blank	0.000	01:53
Subject Sample # 1	0.221	01:54
Breath Volume	2.394	01:54
Air Blank	0.000	01:54
Air Blank	0.000	01:56
Subject Sample # 2	0.211	01:57
Breath Volume	2.414	01:57
Air Blank	0.000	01:57
Control Test	0.080	01:58
Air Blank	0.000	01:58
Diagnostics Check	OK	01:58

Data Download Date/Time: 04/01/2019 07:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: MCLAUGHLIN

DATE/TIME OF TEST: 02/03/2019 01:02:26

Diagnostics Check	OK	00:59	
Air Blank	0.000	00:59	
Control Test	0.080	01:00	
Air Blank	0.000	01:00	
Subject Sample	# 1	01:01	Subject Test Refused
Breath Volume	0.000	01:01	
Air Blank	0.000	01:01	
Control Test	0.079	01:01	
Air Blank	0.000	01:02	
Diagnostics Check	OK	01:02	

Data Download Date/Time: 03/20/2019 22:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICOTTILLI, JOSHUA, A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI, JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: BLOMBERG

DATE/TIME OF TEST: 02/03/2019 02:36:18

Diagnostics Check	OK	02:33	
Air Blank	0.000	02:33	
Control Test	0.080	02:34	
Air Blank	0.000	02:34	
Subject Sample	# 1	02:34	Subject Test Refused
Breath Volume	0.000	02:34	
Air Blank	0.000	02:35	
Control Test	0.080	02:35	
Air Blank	0.000	02:36	
Diagnostics Check	OK	02:36	

Data Download Date/Time: 03/20/2019 22:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BEHRENSHOUSER,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: FAYAD

DATE/TIME OF TEST: 02/03/2019 03:57:49

Diagnostics Check	OK	03:51
Air Blank	0.000	03:51
Control Test	0.080	03:52
Air Blank	0.000	03:52
Subject Sample # 1	0.179	03:53
Breath Volume	1.660	03:53
Air Blank	0.000	03:54
Air Blank	0.000	03:55
Subject Sample # 2	0.192	03:56
Breath Volume	2.886	03:56
Air Blank	0.000	03:56
Control Test	0.079	03:57
Air Blank	0.000	03:57
Diagnostics Check	OK	03:57

Data Download Date/Time: 03/20/2019 22:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEDE,BRENNEN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WEDE,BRENNEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: SANDS

DATE/TIME OF TEST: 02/03/2019 04:53:03

Diagnostics Check	OK	04:46
Air Blank	0.000	04:47
Control Test	0.080	04:47
Air Blank	0.000	04:48
Subject Sample # 1	0.000	04:48
Breath Volume	1.410	04:48
Air Blank	0.000	04:49
Air Blank	0.000	04:51
Subject Sample # 2	0.000	04:51
Breath Volume	1.382	04:51
Air Blank	0.000	04:52
Control Test	0.080	04:52
Air Blank	0.000	04:52
Diagnostics Check	OK	04:53

Data Download Date/Time: 03/20/2019 22:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE,DEPUTY WILLIAM,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WILTSE,DEPUTY WILLIAM
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: LOPEZ ROBLES

DATE/TIME OF TEST: 02/06/2019 03:05:03

Diagnostics Check	OK	02:58
Air Blank	0.000	02:58
Control Test	0.080	02:59
Air Blank	0.000	02:59
Subject Sample # 1	0.138	03:00
Breath Volume	1.234	03:00
Air Blank	0.000	03:01
Air Blank	0.000	03:02
Subject Sample # 2	0.156	03:03
Breath Volume	1.593	03:03
Air Blank	0.000	03:04
Control Test	0.081	03:04
Air Blank	0.000	03:04
Diagnostics Check	OK	03:05

Data Download Date/Time: 03/20/2019 22:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICOTTILLI,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI,JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: LONG

DATE/TIME OF TEST: 02/07/2019 23:40:17

Diagnostics Check	OK	23:30
Air Blank	0.000	23:31
Control Test	0.080	23:31
Air Blank	0.000	23:32
Subject Sample # 1	0.203	23:34
Breath Volume	1.359	23:34
Air Blank	0.000	23:35
Air Blank	0.000	23:36
Subject Sample # 2	0.207	23:38
Breath Volume	2.757	23:38
Air Blank	0.000	23:39
Control Test	0.080	23:39
Air Blank	0.000	23:40
Diagnostics Check	OK	23:40

Data Download Date/Time: 03/20/2019 22:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 02/08/2019 02:41:57

Diagnostics Check	OK	02:30	
Air Blank	0.000	02:31	
Control Test	0.081	02:31	
Air Blank	0.000	02:31	
Subject Sample	# 1	02:35	No Sample Provided
Breath Volume	0.000	02:35	
Air Blank	0.000	02:35	
Air Blank	0.000	02:37	
Subject Sample	# 2	02:40	No Sample Provided
Breath Volume	0.000	02:40	
Air Blank	0.000	02:41	
Control Test	0.082	02:41	
Air Blank	0.000	02:41	
Diagnostics Check	OK	02:41	

Data Download Date/Time: 03/20/2019 22:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE,DEPUTY WILLIAM,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CORSER,DEPUTY WILLIAM
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:41

SUBJECT'S LAST NAME: GERARDE

DATE/TIME OF TEST: 02/08/2019 19:09:35

Diagnostics Check	OK	19:02
Air Blank	0.000	19:02
Control Test	0.081	19:03
Air Blank	0.000	19:03
Subject Sample # 1	0.264	19:04
Breath Volume	1.671	19:04
Air Blank	0.000	19:05
Air Blank	0.000	19:07
Subject Sample # 2	0.277	19:07
Breath Volume	2.097	19:07
Air Blank	0.000	19:08
Control Test	0.081	19:08
Air Blank	0.000	19:09
Diagnostics Check	OK	19:09

Data Download Date/Time: 03/20/2019 22:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICOTTILLI, JOSHUA, A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI, JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: HART

DATE/TIME OF TEST: 02/08/2019 20:58:19

Diagnostics Check	OK	20:50
Air Blank	0.000	20:50
Control Test	0.081	20:50
Air Blank	0.000	20:51
Subject Sample # 1	0.212	20:53
Breath Volume	1.320	20:53
Air Blank	0.000	20:53
Air Blank	0.000	20:55
Subject Sample # 2	0.205	20:56
Breath Volume	1.828	20:56
Air Blank	0.000	20:57
Control Test	0.081	20:57
Air Blank	0.000	20:58
Diagnostics Check	OK	20:58

Data Download Date/Time: 03/20/2019 22:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 02/08/2019 22:28:34

Diagnostics Check	OK	22:22
Air Blank	0.000	22:22
Control Test	0.080	22:23
Air Blank	0.000	22:23
Subject Sample # 1	0.185	22:24
Breath Volume	2.816	22:24
Air Blank	0.000	22:24
Air Blank	0.000	22:26
Subject Sample # 2	0.198	22:27
Breath Volume	2.609	22:27
Air Blank	0.000	22:27
Control Test	0.080	22:27
Air Blank	0.000	22:28
Diagnostics Check	OK	22:28

Data Download Date/Time: 03/20/2019 22:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICOTTILLI, JOSHUA, A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI, JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: DIPASCALE

DATE/TIME OF TEST: 02/09/2019 01:48:03

Diagnostics Check	OK	01:38
Air Blank	0.000	01:39
Control Test	0.079	01:39
Air Blank	0.000	01:40
Subject Sample # 1	0.187	01:42
Breath Volume	1.105	01:42
Air Blank	0.000	01:43
Air Blank	0.000	01:44
Subject Sample # 2	0.202	01:46
Breath Volume	1.253	01:46
Air Blank	0.000	01:47
Control Test	0.079	01:47
Air Blank	0.000	01:47
Diagnostics Check	OK	01:48

Data Download Date/Time: 03/20/2019 22:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE,DEPUTY WILLIAM,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WILTSE,DEPUTY WILLIAM
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: STOKES

DATE/TIME OF TEST: 02/09/2019 04:27:40

Diagnostics Check	OK	04:21
Air Blank	0.000	04:21
Control Test	0.080	04:22
Air Blank	0.000	04:22
Subject Sample # 1	0.112	04:23
Breath Volume	1.394	04:23
Air Blank	0.000	04:23
Air Blank	0.000	04:25
Subject Sample # 2	0.109	04:26
Breath Volume	1.421	04:26
Air Blank	0.000	04:26
Control Test	0.080	04:27
Air Blank	0.000	04:27
Diagnostics Check	OK	04:27

Data Download Date/Time: 03/20/2019 22:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RASSIER,TOMMY,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SMITH,ROBERT
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:52

SUBJECT'S LAST NAME: BAK

DATE/TIME OF TEST: 02/09/2019 16:16:46

Diagnostics Check	OK	16:13	
Air Blank	0.000	16:14	
Control Test	0.080	16:14	
Air Blank	0.000	16:15	
Subject Sample	# 1	16:15	Subject Test Refused
Breath Volume	0.000	16:15	
Air Blank	0.000	16:15	
Control Test	0.080	16:16	
Air Blank	0.000	16:16	
Diagnostics Check	OK	16:16	

Data Download Date/Time: 03/20/2019 22:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICOTTILLI, JOSHUA, A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI, JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:17

SUBJECT'S LAST NAME: SIDAROUS

DATE/TIME OF TEST: 02/09/2019 20:40:56

Diagnostics Check	OK	20:37	
Air Blank	0.000	20:37	
Control Test	0.079	20:38	
Air Blank	0.000	20:38	
Subject Sample	# 1	20:39	Subject Test Refused
Breath Volume	0.000	20:39	
Air Blank	0.000	20:40	
Control Test	0.081	20:40	
Air Blank	0.000	20:40	
Diagnostics Check	OK	20:40	

Data Download Date/Time: 03/20/2019 22:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: BAUER

DATE/TIME OF TEST: 02/09/2019 21:36:02

Diagnostics Check	OK	21:32	
Air Blank	0.000	21:33	
Control Test	0.081	21:33	
Air Blank	0.000	21:34	
Subject Sample	# 1	21:34	Subject Test Refused
Breath Volume	0.000	21:34	
Air Blank	0.000	21:35	
Control Test	0.081	21:35	
Air Blank	0.000	21:35	
Diagnostics Check	OK	21:36	

Data Download Date/Time: 03/20/2019 22:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: SNOW

DATE/TIME OF TEST: 02/09/2019 23:35:51

Diagnostics Check	OK	23:29
Air Blank	0.000	23:30
Control Test	0.081	23:30
Air Blank	0.000	23:31
Subject Sample # 1	0.128	23:31
Breath Volume	1.761	23:31
Air Blank	0.000	23:32
Air Blank	0.000	23:33
Subject Sample # 2	0.123	23:34
Breath Volume	1.472	23:34
Air Blank	0.000	23:34
Control Test	0.080	23:35
Air Blank	0.000	23:35
Diagnostics Check	OK	23:35

Data Download Date/Time: 03/20/2019 22:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: DIEHL

DATE/TIME OF TEST: 02/10/2019 01:45:33

Diagnostics Check	OK	01:34	
Air Blank	0.000	01:34	
Control Test	0.081	01:34	
Air Blank	0.000	01:35	
Subject Sample	# 1	01:38	No Sample Provided
Breath Volume	0.000	01:38	
Air Blank	0.000	01:39	
Air Blank	0.000	01:41	
Subject Sample	# 2	01:44	No Sample Provided
Breath Volume	0.000	01:44	
Air Blank	0.000	01:44	
Control Test	0.080	01:44	
Air Blank	0.000	01:45	
Diagnostics Check	OK	01:45	

Data Download Date/Time: 03/20/2019 22:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE,DEPUTY WILLIAM,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CASTELEIRO,DEPUTY A.
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: DUBOIS

DATE/TIME OF TEST: 02/10/2019 20:19:24

Diagnostics Check	OK	20:13
Air Blank	0.000	20:13
Control Test	0.081	20:13
Air Blank	0.000	20:14
Subject Sample # 1	0.201	20:14
Breath Volume	1.566	20:14
Air Blank	0.000	20:15
Air Blank	0.000	20:17
Subject Sample # 2	0.219	20:17
Breath Volume	1.554	20:17
Air Blank	0.000	20:18
Control Test	0.080	20:18
Air Blank	0.000	20:19
Diagnostics Check	OK	20:19

Data Download Date/Time: 03/20/2019 22:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONGMIRE,BENJAMIN,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: LONGMIRE,BENJAMIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: ODUM

DATE/TIME OF TEST: 02/10/2019 21:42:22

Diagnostics Check	OK	21:36
Air Blank	0.000	21:36
Control Test	0.081	21:36
Air Blank	0.000	21:37
Subject Sample # 1	0.182	21:37
Breath Volume	2.941	21:37
Air Blank	0.000	21:38
Air Blank	0.000	21:40
Subject Sample # 2	0.180	21:40
Breath Volume	2.789	21:40
Air Blank	0.000	21:41
Control Test	0.080	21:41
Air Blank	0.000	21:42
Diagnostics Check	OK	21:42

Data Download Date/Time: 03/20/2019 22:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE,DEPUTY WILLIAM,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,DEPUTY L.
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: QUACKENBOS

DATE/TIME OF TEST: 02/10/2019 23:21:02

Diagnostics Check	OK	23:14
Air Blank	0.000	23:15
Control Test	0.081	23:15
Air Blank	0.000	23:16
Subject Sample # 1	0.000	23:16
Breath Volume	3.093	23:16
Air Blank	0.000	23:17
Air Blank	0.000	23:19
Subject Sample # 2	0.000	23:19
Breath Volume	2.589	23:19
Air Blank	0.000	23:20
Control Test	0.081	23:20
Air Blank	0.000	23:20
Diagnostics Check	OK	23:21

Data Download Date/Time: 03/20/2019 22:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE,DEPUTY WILLIAM,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DUCKERS,DEPUTY R.
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: MCPEEK

DATE/TIME OF TEST: 02/10/2019 23:53:22

Diagnostics Check	OK	23:50	
Air Blank	0.000	23:50	
Control Test	0.081	23:51	
Air Blank	0.000	23:51	
Subject Sample	# 1	23:52	Subject Test Refused
Breath Volume	0.000	23:52	
Air Blank	0.000	23:52	
Control Test	0.081	23:52	
Air Blank	0.000	23:53	
Diagnostics Check	OK	23:53	

Data Download Date/Time: 03/20/2019 22:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICOTTILLI, JOSHUA, A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI, JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: FORBES

DATE/TIME OF TEST: 02/14/2019 01:52:13

Diagnostics Check	OK	01:45
Air Blank	0.000	01:45
Control Test	0.081	01:45
Air Blank	0.000	01:46
Subject Sample # 1	0.019	01:47
Breath Volume	1.554	01:47
Air Blank	0.000	01:47
Air Blank	0.000	01:49
Subject Sample # 2	0.016	01:50
Breath Volume	1.601	01:50
Air Blank	0.000	01:51
Control Test	0.081	01:51
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

Data Download Date/Time: 03/20/2019 22:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: MARROCCO

DATE/TIME OF TEST: 02/15/2019 03:14:53

Diagnostics Check	OK	03:08
Air Blank	0.000	03:09
Control Test	0.080	03:09
Air Blank	0.000	03:09
Subject Sample # 1	0.082	03:10
Breath Volume	3.582	03:10
Air Blank	0.000	03:10
Air Blank	0.000	03:12
Subject Sample # 2	0.085	03:13
Breath Volume	3.625	03:13
Air Blank	0.000	03:13
Control Test	0.081	03:14
Air Blank	0.000	03:14
Diagnostics Check	OK	03:14

Data Download Date/Time: 03/20/2019 22:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: LAMKIN

DATE/TIME OF TEST: 02/15/2019 04:20:52

Diagnostics Check	OK	04:14
Air Blank	0.000	04:15
Control Test	0.081	04:15
Air Blank	0.000	04:16
Subject Sample # 1	0.065	04:16
Breath Volume	1.527	04:16
Air Blank	0.000	04:17
Air Blank	0.000	04:18
Subject Sample # 2	0.068	04:19
Breath Volume	1.531	04:19
Air Blank	0.000	04:19
Control Test	0.080	04:20
Air Blank	0.000	04:20
Diagnostics Check	OK	04:20

Data Download Date/Time: 03/20/2019 22:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: WARD

DATE/TIME OF TEST: 02/16/2019 02:52:11

Diagnostics Check	OK	02:40	
Air Blank	0.000	02:41	
Control Test	0.080	02:41	
Air Blank	0.000	02:42	
Subject Sample	# 1	02:45	No Sample Provided
Breath Volume	0.000	02:45	
Air Blank	0.000	02:45	
Air Blank	0.000	02:47	
Subject Sample	# 2	02:50	No Sample Provided
Breath Volume	0.000	02:50	
Air Blank	0.000	02:51	
Control Test	0.081	02:51	
Air Blank	0.000	02:52	
Diagnostics Check	OK	02:52	

Data Download Date/Time: 03/20/2019 22:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE,DEPUTY WILLIAM,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,DEPUTY L.
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 02/19/2019 02:36:08

Diagnostics Check	OK	02:33	
Air Blank	0.000	02:33	
Control Test	0.081	02:33	
Air Blank	0.000	02:34	
Subject Sample	# 1	02:34	Subject Test Refused
Breath Volume	0.000	02:34	
Air Blank	0.000	02:35	
Control Test	0.081	02:35	
Air Blank	0.000	02:36	
Diagnostics Check	OK	02:36	

Data Download Date/Time: 03/20/2019 22:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE,DEPUTY WILLIAM,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SKALCO,DEPUTY RYAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: STENSTROM

DATE/TIME OF TEST: 02/20/2019 03:39:47

Diagnostics Check	OK	03:32
Air Blank	0.000	03:33
Control Test	0.081	03:33
Air Blank	0.000	03:34
Subject Sample # 1	0.196	03:34
Breath Volume	4.101	03:34
Air Blank	0.000	03:35
Air Blank	0.000	03:37
Subject Sample # 2	0.194	03:38
Breath Volume	2.140	03:38
Air Blank	0.000	03:38
Control Test	0.081	03:39
Air Blank	0.000	03:39
Diagnostics Check	OK	03:39

Data Download Date/Time: 03/20/2019 22:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE,DEPUTY WILLIAM,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,DEPUTY L.
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: CUBE

DATE/TIME OF TEST: 02/20/2019 04:45:01

Diagnostics Check	OK	04:38
Air Blank	0.000	04:38
Control Test	0.080	04:39
Air Blank	0.000	04:39
Subject Sample # 1	0.237	04:40
Breath Volume	2.175	04:40
Air Blank	0.000	04:41
Air Blank	0.000	04:43
Subject Sample # 2	0.239	04:43
Breath Volume	2.156	04:43
Air Blank	0.000	04:44
Control Test	0.081	04:44
Air Blank	0.000	04:44
Diagnostics Check	OK	04:45

Data Download Date/Time: 03/20/2019 22:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: VESELINOVIC

DATE/TIME OF TEST: 02/21/2019 01:06:39

Diagnostics Check	OK	01:00
Air Blank	0.000	01:00
Control Test	0.080	01:01
Air Blank	0.000	01:01
Subject Sample # 1	0.169	01:02
Breath Volume	1.277	01:02
Air Blank	0.000	01:02
Air Blank	0.000	01:04
Subject Sample # 2	0.178	01:05
Breath Volume	1.546	01:05
Air Blank	0.000	01:05
Control Test	0.081	01:06
Air Blank	0.000	01:06
Diagnostics Check	OK	01:06

Data Download Date/Time: 03/20/2019 22:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICOTTILLI,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI,JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: MELECCI

DATE/TIME OF TEST: 02/22/2019 03:02:30

Diagnostics Check	OK	02:59	
Air Blank	0.000	03:00	
Control Test	0.080	03:00	
Air Blank	0.000	03:00	
Subject Sample	# 1	03:01	Subject Test Refused
Breath Volume	0.000	03:01	
Air Blank	0.000	03:01	
Control Test	0.080	03:01	
Air Blank	0.000	03:02	
Diagnostics Check	OK	03:02	

Data Download Date/Time: 03/20/2019 22:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: ROUPE

DATE/TIME OF TEST: 02/22/2019 04:54:46

Diagnostics Check	OK	04:51	
Air Blank	0.000	04:52	
Control Test	0.081	04:52	
Air Blank	0.000	04:53	
Subject Sample	# 1	04:53	Subject Test Refused
Breath Volume	0.000	04:53	
Air Blank	0.000	04:53	
Control Test	0.080	04:54	
Air Blank	0.000	04:54	
Diagnostics Check	OK	04:54	

Data Download Date/Time: 03/20/2019 22:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHORT, WILLIAM, C
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: NETTLES, JIMMY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 14:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:31

SUBJECT'S LAST NAME: ANDREWS

DATE/TIME OF TEST: 02/22/2019 15:58:55

Diagnostics Check	OK	15:52
Air Blank	0.000	15:52
Control Test	0.080	15:53
Air Blank	0.000	15:53
Subject Sample # 1	0.000	15:54
Breath Volume	1.621	15:54
Air Blank	0.000	15:54
Air Blank	0.000	15:56
Subject Sample # 2	0.000	15:57
Breath Volume	1.527	15:57
Air Blank	0.000	15:58
Control Test	0.080	15:58
Air Blank	0.000	15:58
Diagnostics Check	OK	15:58

Data Download Date/Time: 03/20/2019 22:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICOTTILLI, JOSHUA, A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI, JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: PEREIRA

DATE/TIME OF TEST: 02/23/2019 01:07:32

Diagnostics Check	OK	01:00
Air Blank	0.000	01:00
Control Test	0.081	01:01
Air Blank	0.000	01:01
Subject Sample # 1	0.103	01:02
Breath Volume	1.621	01:02
Air Blank	0.000	01:03
Air Blank	0.000	01:04
Subject Sample # 2	0.104	01:06
Breath Volume	1.214	01:06
Air Blank	0.000	01:06
Control Test	0.081	01:06
Air Blank	0.000	01:07
Diagnostics Check	OK	01:07

Data Download Date/Time: 03/20/2019 22:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, WILLIAM, T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: CRESPO

DATE/TIME OF TEST: 02/23/2019 02:24:06

Diagnostics Check	OK	02:16
Air Blank	0.000	02:17
Control Test	0.082	02:17
Air Blank	0.000	02:17
Subject Sample # 1	0.052	02:19
Breath Volume	2.304	02:19
Air Blank	0.000	02:19
Air Blank	0.000	02:21
Subject Sample # 2	0.053	02:22
Breath Volume	2.929	02:22
Air Blank	0.000	02:23
Control Test	0.081	02:23
Air Blank	0.000	02:24
Diagnostics Check	OK	02:24

Data Download Date/Time: 03/20/2019 22:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE,DEPUTY WILLIAM,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WILTSE,DEPUTY WILLIAM
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: HENDRIX

DATE/TIME OF TEST: 02/23/2019 04:03:55

Diagnostics Check	OK	04:01	
Air Blank	0.000	04:01	
Control Test	0.080	04:01	
Air Blank	0.000	04:02	
Subject Sample	# 1	04:02	Subject Test Refused
Breath Volume	0.000	04:02	
Air Blank	0.000	04:03	
Control Test	0.081	04:03	
Air Blank	0.000	04:03	
Diagnostics Check	OK	04:03	

Data Download Date/Time: 03/20/2019 22:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICOTTILLI, JOSHUA, A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI, JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: JAMES

DATE/TIME OF TEST: 02/24/2019 01:19:18

Diagnostics Check	OK	01:10
Air Blank	0.000	01:11
Control Test	0.081	01:11
Air Blank	0.000	01:12
Subject Sample # 1	0.139	01:14
Breath Volume	1.667	01:14
Air Blank	0.000	01:14
Air Blank	0.000	01:16
Subject Sample # 2	0.141	01:17
Breath Volume	1.777	01:17
Air Blank	0.000	01:18
Control Test	0.080	01:18
Air Blank	0.000	01:19
Diagnostics Check	OK	01:19

Data Download Date/Time: 03/20/2019 22:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE,DEPUTY WILLIAM,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WILTSE,DEPUTY WILLIAM
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: CLARIT

DATE/TIME OF TEST: 02/24/2019 01:58:25

Diagnostics Check	OK	01:55	
Air Blank	0.000	01:55	
Control Test	0.081	01:56	
Air Blank	0.000	01:56	
Subject Sample	# 1	01:57	Subject Test Refused
Breath Volume	0.000	01:57	
Air Blank	0.000	01:57	
Control Test	0.081	01:57	
Air Blank	0.000	01:58	
Diagnostics Check	OK	01:58	

Data Download Date/Time: 03/20/2019 22:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEDE,BRENNEN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WEDE,BRENNEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: NONTHE MARTINEZ

DATE/TIME OF TEST: 02/24/2019 02:38:42

Diagnostics Check	OK	02:32
Air Blank	0.000	02:32
Control Test	0.080	02:33
Air Blank	0.000	02:33
Subject Sample # 1	0.162	02:34
Breath Volume	1.117	02:34
Air Blank	0.000	02:34
Air Blank	0.000	02:36
Subject Sample # 2	0.159	02:37
Breath Volume	1.371	02:37
Air Blank	0.000	02:37
Control Test	0.079	02:38
Air Blank	0.000	02:38
Diagnostics Check	OK	02:38

Data Download Date/Time: 03/20/2019 22:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: MOTA

DATE/TIME OF TEST: 02/24/2019 04:50:25

Diagnostics Check	OK	04:46	
Air Blank	0.000	04:47	
Control Test	0.082	04:47	
Air Blank	0.000	04:48	
Subject Sample	# 1	04:49	Subject Test Refused
Breath Volume	0.000	04:49	
Air Blank	0.000	04:49	
Control Test	0.081	04:49	
Air Blank	0.000	04:50	
Diagnostics Check	OK	04:50	

Data Download Date/Time: 03/20/2019 22:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, WILLIAM, T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 02/25/2019 00:18:52

Diagnostics Check	OK	00:14	
Air Blank	0.000	00:15	
Control Test	0.081	00:15	
Air Blank	0.000	00:15	
Subject Sample	# 1	00:17	Subject Test Refused
Breath Volume	0.000	00:17	
Air Blank	0.000	00:17	
Control Test	0.081	00:18	
Air Blank	0.000	00:18	
Diagnostics Check	OK	00:18	

Data Download Date/Time: 03/20/2019 22:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEDE,BRENNEN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MUNOZ,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 02/25/2019 04:37:42

Diagnostics Check	OK	04:31
Air Blank	0.000	04:32
Control Test	0.081	04:32
Air Blank	0.000	04:32
Subject Sample # 1	0.281	04:33
Breath Volume	1.519	04:33
Air Blank	0.000	04:33
Air Blank	0.000	04:35
Subject Sample # 2	0.282	04:36
Breath Volume	1.539	04:36
Air Blank	0.000	04:36
Control Test	0.080	04:37
Air Blank	0.000	04:37
Diagnostics Check	OK	04:37

Data Download Date/Time: 03/20/2019 22:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORNING,PETER,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DUNNING,RYAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:46

SUBJECT'S LAST NAME: CREECH

DATE/TIME OF TEST: 02/25/2019 18:09:25

Diagnostics Check	OK	18:06	
Air Blank	0.000	18:07	
Control Test	0.082	18:07	
Air Blank	0.000	18:07	
Subject Sample	# 1	18:08	Subject Test Refused
Breath Volume	0.000	18:08	
Air Blank	0.000	18:08	
Control Test	0.081	18:08	
Air Blank	0.000	18:09	
Diagnostics Check	OK	18:09	

Data Download Date/Time: 03/20/2019 22:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNAVELY,KENT,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: STILES,SHANE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 11:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:42

SUBJECT'S LAST NAME: LEGERE

DATE/TIME OF TEST: 02/04/2019 13:09:22

Diagnostics Check	OK	13:03
Air Blank	0.000	13:03
Control Test	0.079	13:03
Air Blank	0.000	13:04
Subject Sample # 1	0.240	13:04
Breath Volume	3.519	13:04
Air Blank	0.000	13:05
Air Blank	0.000	13:07
Subject Sample # 2	0.240	13:07
Breath Volume	2.800	13:07
Air Blank	0.000	13:08
Control Test	0.078	13:08
Air Blank	0.000	13:09
Diagnostics Check	OK	13:09

Data Download Date/Time: 03/20/2019 22:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MUNOZ,B
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: EASON

DATE/TIME OF TEST: 02/09/2019 04:57:40

Diagnostics Check	OK	04:50
Air Blank	0.000	04:50
Control Test	0.080	04:51
Air Blank	0.000	04:51
Subject Sample # 1	0.189	04:52
Breath Volume	3.085	04:52
Air Blank	0.000	04:53
Air Blank	0.000	04:54
Subject Sample # 2	0.179	04:55
Breath Volume	3.066	04:55
Air Blank	0.000	04:56
Control Test	0.080	04:57
Air Blank	0.000	04:57
Diagnostics Check	OK	04:57

Data Download Date/Time: 03/20/2019 22:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FRASER,CHRISTOPHER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: MELENDEZ CLARK

DATE/TIME OF TEST: 02/11/2019 04:20:03

Diagnostics Check	OK	04:17	
Air Blank	0.000	04:17	
Control Test	0.080	04:17	
Air Blank	0.000	04:18	
Subject Sample	# 1	04:18	Subject Test Refused
Breath Volume	0.000	04:18	
Air Blank	0.000	04:19	
Control Test	0.080	04:19	
Air Blank	0.000	04:19	
Diagnostics Check	OK	04:20	

Data Download Date/Time: 03/20/2019 22:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: REIDY,TYLER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: PILIPAUSKAS

DATE/TIME OF TEST: 02/11/2019 05:01:32

Diagnostics Check	OK	04:53
Air Blank	0.000	04:54
Control Test	0.080	04:54
Air Blank	0.000	04:55
Subject Sample # 1	0.154	04:56
Breath Volume	2.042	04:56
Air Blank	0.000	04:56
Air Blank	0.000	04:58
Subject Sample # 2	0.146	04:59
Breath Volume	2.125	04:59
Air Blank	0.000	05:00
Control Test	0.079	05:00
Air Blank	0.000	05:01
Diagnostics Check	OK	05:01

Data Download Date/Time: 03/20/2019 22:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DAVIS,DEPUTY J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 14:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:30

SUBJECT'S LAST NAME: OVERMAN

DATE/TIME OF TEST: 02/11/2019 14:57:53

Diagnostics Check	OK	14:50
Air Blank	0.000	14:51
Control Test	0.079	14:51
Air Blank	0.000	14:51
Subject Sample # 1	0.000	14:52
Breath Volume	1.519	14:52
Air Blank	0.000	14:52
Air Blank	0.000	14:54
Subject Sample # 2	0.000	14:56
Breath Volume	1.570	14:56
Air Blank	0.000	14:56
Control Test	0.080	14:57
Air Blank	0.000	14:57
Diagnostics Check	OK	14:57

Data Download Date/Time: 03/20/2019 22:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVARTS,JASON,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: EVARTS,JASON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:36

SUBJECT'S LAST NAME: COWLEY

DATE/TIME OF TEST: 02/14/2019 21:43:49

Diagnostics Check	OK	21:41	
Air Blank	0.000	21:42	
Control Test	0.080	21:42	
Air Blank	0.000	21:43	
Subject Sample	# 1	21:43	Sequence Aborted
Breath Volume	0.000	21:43	
Air Blank	0.000	21:43	

Data Download Date/Time: 03/20/2019 22:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVARTS,JASON,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: EVARTS,JASON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: COWLEY

DATE/TIME OF TEST: 02/14/2019 22:01:32

Diagnostics Check	OK	21:57	
Air Blank	0.000	21:57	
Control Test	0.080	21:58	
Air Blank	0.000	21:58	
Subject Sample	# 1	22:00	Subject Test Refused
Breath Volume	0.000	22:00	
Air Blank	0.000	22:00	
Control Test	0.080	22:00	
Air Blank	0.000	22:01	
Diagnostics Check	OK	22:01	

Data Download Date/Time: 03/20/2019 22:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: HAWKINS

DATE/TIME OF TEST: 02/16/2019 03:22:43

Diagnostics Check	OK	03:16
Air Blank	0.000	03:16
Control Test	0.079	03:17
Air Blank	0.000	03:17
Subject Sample # 1	0.155	03:18
Breath Volume	1.304	03:18
Air Blank	0.000	03:19
Air Blank	0.000	03:20
Subject Sample # 2	0.152	03:21
Breath Volume	1.894	03:21
Air Blank	0.000	03:21
Control Test	0.079	03:22
Air Blank	0.000	03:22
Diagnostics Check	OK	03:22

Data Download Date/Time: 03/20/2019 22:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNAVELY,KENT,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HANEWINKEL,BIANCA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 09:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:36

SUBJECT'S LAST NAME: FIELDS

DATE/TIME OF TEST: 02/19/2019 11:03:05

Diagnostics Check	OK	10:56
Air Blank	0.000	10:57
Control Test	0.080	10:57
Air Blank	0.000	10:58
Subject Sample # 1	0.171	10:58
Breath Volume	1.871	10:58
Air Blank	0.000	10:59
Air Blank	0.000	11:00
Subject Sample # 2	0.176	11:01
Breath Volume	2.074	11:01
Air Blank	0.000	11:02
Control Test	0.080	11:02
Air Blank	0.000	11:02
Diagnostics Check	OK	11:03

Data Download Date/Time: 03/20/2019 22:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RANISZEWSKI,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 02/24/2019 05:00:00

Diagnostics Check	OK	04:52
Air Blank	0.000	04:53
Control Test	0.079	04:53
Air Blank	0.000	04:54
Subject Sample # 1	0.196	04:55
Breath Volume	1.972	04:55
Air Blank	0.000	04:56
Air Blank	0.000	04:57
Subject Sample # 2	0.205	04:58
Breath Volume	2.058	04:58
Air Blank	0.000	04:59
Control Test	0.079	04:59
Air Blank	0.000	04:59
Diagnostics Check	OK	05:00

Data Download Date/Time: 03/20/2019 22:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

INSTRUMENT REGISTERED TO: INDIALANTIC P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOURNIER,TIMOTHY,R
BREATH TEST OPERATOR AGENCY: Indialantic PD
ARRESTING OFFICER: COCCHIOLA,AUSTIN
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 01/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: CALLENDER

DATE/TIME OF TEST: 02/09/2019 02:20:00

Diagnostics Check	OK	02:12
Air Blank	0.000	02:13
Control Test	0.080	02:13
Air Blank	0.000	02:13
Subject Sample # 1	0.183	02:15
Breath Volume	2.343	02:15
Air Blank	0.000	02:16
Air Blank	0.000	02:17
Subject Sample # 2	0.171	02:18
Breath Volume	1.683	02:18
Air Blank	0.000	02:19
Control Test	0.078	02:19
Air Blank	0.000	02:19
Diagnostics Check	OK	02:20

Data Download Date/Time: 02/27/2019 20:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

INSTRUMENT REGISTERED TO: INDIALANTIC P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLSTINE,SCOTT,
BREATH TEST OPERATOR AGENCY: Indialantic PD
ARRESTING OFFICER: PARRISH,JASON
ARRESTING OFFICER AGENCY: Indialantic PD
LAST AGENCY INSPECTION: 01/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: SHIMP

DATE/TIME OF TEST: 02/10/2019 00:52:10

Diagnostics Check	OK	00:40	
Air Blank	0.000	00:41	
Control Test	0.080	00:41	
Air Blank	0.000	00:42	
Subject Sample	# 1	00:45	No Sample Provided
Breath Volume	0.000	00:45	
Air Blank	0.000	00:45	
Air Blank	0.000	00:47	
Subject Sample	# 2	00:50	No Sample Provided
Breath Volume	0.000	00:50	
Air Blank	0.000	00:51	
Control Test	0.079	00:51	
Air Blank	0.000	00:52	
Diagnostics Check	OK	00:52	

Data Download Date/Time: 02/27/2019 20:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

INSTRUMENT REGISTERED TO: INDIALANTIC P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,TIMOTHY,C
BREATH TEST OPERATOR AGENCY: Indialantic PD
ARRESTING OFFICER: WEBER,TIMOTHY
ARRESTING OFFICER AGENCY: Indialantic PD
LAST AGENCY INSPECTION: 01/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: HASKOOR

DATE/TIME OF TEST: 02/24/2019 01:10:16

Diagnostics Check	OK	01:03
Air Blank	0.000	01:03
Control Test	0.079	01:04
Air Blank	0.000	01:04
Subject Sample # 1	0.257	01:05
Breath Volume	1.335	01:05
Air Blank	0.000	01:06
Air Blank	0.000	01:08
Subject Sample # 2	0.277	01:08
Breath Volume	1.324	01:08
Air Blank	0.000	01:09
Control Test	0.079	01:09
Air Blank	0.000	01:10
Diagnostics Check	OK	01:10

Data Download Date/Time: 02/27/2019 20:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORENZO,BYRON,C
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: ROGERS,DWIGHT
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 01/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: LASKETT

DATE/TIME OF TEST: 02/16/2019 02:24:41

Diagnostics Check	OK	02:24	
Air Blank		02:24	Improper Sample
Air Blank		02:24	Improper Sample

Data Download Date/Time: 02/20/2019 06:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORENZO,BYRON,C
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: ROGERS,DWIGHT
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 01/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: LASKETT

DATE/TIME OF TEST: 02/16/2019 02:27:38

Diagnostics Check	OK	02:27	
Air Blank		02:27	Improper Sample
Air Blank		02:27	Improper Sample

Data Download Date/Time: 02/20/2019 06:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORENZO,BYRON,C
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: ROGERS,DWIGHT
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 01/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: LASKETT

DATE/TIME OF TEST: 02/16/2019 02:30:41

Diagnostics Check	OK	02:30	
Air Blank		02:30	Improper Sample
Air Blank		02:30	Improper Sample

Data Download Date/Time: 02/20/2019 06:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORENZO,BYRON,C
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: ROGERS,DWIGHT
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 01/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: LASKETT

DATE/TIME OF TEST: 02/16/2019 02:39:00

Diagnostics Check	OK	02:38	
Air Blank		02:38	Ambient Fail
Air Blank	0.000	02:39	

Data Download Date/Time: 02/20/2019 06:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORENZO,BYRON,C
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: ROGERS,DWIGHT
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 01/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: LASKETT

DATE/TIME OF TEST: 02/16/2019 02:48:02

Diagnostics Check	OK	02:41
Air Blank	0.000	02:41
Control Test	0.080	02:42
Air Blank	0.000	02:42
Subject Sample # 1	0.148	02:43
Breath Volume	1.511	02:43
Air Blank	0.000	02:44
Air Blank	0.000	02:45
Subject Sample # 2	0.144	02:46
Breath Volume	2.109	02:46
Air Blank	0.000	02:47
Control Test	0.078	02:47
Air Blank	0.000	02:47
Diagnostics Check	OK	02:48

Data Download Date/Time: 02/20/2019 06:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORENZO,BYRON,C
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: LORENZO,BYRON
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 01/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 02/17/2019 02:21:15

Diagnostics Check	OK	02:13
Air Blank	0.000	02:13
Control Test	0.081	02:14
Air Blank	0.000	02:14
Subject Sample # 1	0.173	02:15
Breath Volume	1.375	02:15
Air Blank	0.000	02:16
Air Blank	0.000	02:18
Subject Sample # 2	0.167	02:19
Breath Volume	1.871	02:19
Air Blank	0.000	02:20
Control Test	0.083	02:20
Air Blank	0.000	02:21
Diagnostics Check	OK	02:21

Data Download Date/Time: 02/20/2019 06:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKSON,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: JACKSON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 01/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 16:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:47

SUBJECT'S LAST NAME: CATON

DATE/TIME OF TEST: 02/17/2019 17:28:00

Diagnostics Check	OK	17:25	
Air Blank	0.000	17:25	
Control Test	0.082	17:25	
Air Blank	0.000	17:26	
Subject Sample	# 1	17:26	Subject Test Refused
Breath Volume	0.000	17:26	
Air Blank	0.000	17:27	
Control Test	0.082	17:27	
Air Blank	0.000	17:27	
Diagnostics Check	OK	17:28	

Data Download Date/Time: 02/20/2019 06:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKSON,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: JACKSON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 01/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 18:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: ROBBINS

DATE/TIME OF TEST: 02/18/2019 18:56:50

Diagnostics Check	OK	18:54	
Air Blank	0.000	18:54	
Control Test	0.082	18:54	
Air Blank	0.000	18:55	
Subject Sample	# 1	18:55	Subject Test Refused
Breath Volume	0.000	18:55	
Air Blank	0.000	18:55	
Control Test	0.081	18:56	
Air Blank	0.000	18:56	
Diagnostics Check	OK	18:56	

Data Download Date/Time: 02/20/2019 06:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HICKS,WILLIAM,L
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: LIGUORI,BRIAN
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 01/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: LOWE

DATE/TIME OF TEST: 02/03/2019 02:53:22

Diagnostics Check	OK	02:46
Air Blank	0.000	02:47
Control Test	0.081	02:47
Air Blank	0.000	02:48
Subject Sample # 1	0.202	02:48
Breath Volume	2.554	02:48
Air Blank	0.000	02:49
Air Blank	0.000	02:51
Subject Sample # 2	0.184	02:51
Breath Volume	2.617	02:51
Air Blank	0.000	02:52
Control Test	0.080	02:52
Air Blank	0.000	02:53
Diagnostics Check	OK	02:53

Data Download Date/Time: 02/26/2019 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LACEY,JOHN,J
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: LACEY,JOHN
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 01/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: HAWKINS

DATE/TIME OF TEST: 02/04/2019 19:40:14

Diagnostics Check	OK	19:30
Air Blank	0.000	19:30
Control Test	0.080	19:31
Air Blank	0.000	19:31
Subject Sample # 1	0.232	19:32
Breath Volume	1.277	19:32
Air Blank	0.000	19:33
Air Blank	0.000	19:35
Subject Sample # 2	0.219	19:38
Breath Volume	1.132	19:38
Air Blank	0.000	19:39
Control Test	0.077	19:39
Air Blank	0.000	19:40
Diagnostics Check	OK	19:40

Data Download Date/Time: 02/26/2019 16:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LACEY,JOHN,J
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: LACEY,JOHN
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 01/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: WELLS

DATE/TIME OF TEST: 02/23/2019 22:45:11

Diagnostics Check	OK	22:35
Air Blank	0.000	22:36
Control Test	0.080	22:36
Air Blank	0.000	22:36
Subject Sample # 1	0.308	22:37
Breath Volume	1.488	22:37
Air Blank	0.000	22:38
Air Blank	0.000	22:39
Subject Sample # 2	0.282	22:40
Breath Volume	2.328	22:40
Air Blank	0.000	22:41
Air Blank	0.000	22:43
Subject Sample # 3	0.291	22:43
Breath Volume	2.574	22:43
Air Blank	0.000	22:44
Control Test	0.078	22:44
Air Blank	0.000	22:45
Diagnostics Check	OK	22:45

Data Download Date/Time: 02/26/2019 16:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LACEY,JOHN,J
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: LACEY,JOHN
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 01/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: DORSAINVIL

DATE/TIME OF TEST: 02/25/2019 03:19:07

Diagnostics Check	OK	03:12
Air Blank	0.000	03:13
Control Test	0.082	03:13
Air Blank	0.000	03:14
Subject Sample # 1	0.140	03:14
Breath Volume	2.707	03:14
Air Blank	0.000	03:15
Air Blank	0.000	03:17
Subject Sample # 2	0.139	03:17
Breath Volume	3.882	03:17
Air Blank	0.000	03:18
Control Test	0.080	03:18
Air Blank	0.000	03:19
Diagnostics Check	OK	03:19

Data Download Date/Time: 02/26/2019 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE, J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 02/01/2019 23:59:55

Diagnostics Check	OK	23:53
Air Blank	0.000	23:53
Control Test	0.083	23:53
Air Blank	0.000	23:54
Subject Sample # 1	0.180	23:55
Breath Volume	1.621	23:55
Air Blank	0.000	23:55
Air Blank	0.000	23:57
Subject Sample # 2	0.176	23:58
Breath Volume	1.562	23:58
Air Blank	0.000	23:58
Control Test	0.082	23:59
Air Blank	0.000	23:59
Diagnostics Check	OK	23:59

Data Download Date/Time: 02/07/2019 13:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN, D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: TOMALEWSKI

DATE/TIME OF TEST: 02/02/2019 01:05:40

Diagnostics Check	OK	00:53	
Air Blank	0.000	00:53	
Control Test	0.082	00:54	
Air Blank	0.000	00:54	
Subject Sample	# 1	00:57	No Sample Provided
Breath Volume	0.000	00:57	
Air Blank	0.000	00:58	
Air Blank	0.000	01:00	
Subject Sample	# 2	01:00	
Breath Volume	2.144	01:00	
Air Blank	0.000	01:01	
Air Blank	0.000	01:03	
Subject Sample	# 3	01:03	
Breath Volume	1.890	01:03	
Air Blank	0.000	01:04	
Control Test	0.082	01:05	
Air Blank	0.000	01:05	
Diagnostics Check	OK	01:05	

Data Download Date/Time: 02/07/2019 13:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: MERCER

DATE/TIME OF TEST: 02/02/2019 02:18:11

Diagnostics Check	OK	02:15	
Air Blank	0.000	02:15	
Control Test	0.083	02:15	
Air Blank	0.000	02:16	
Subject Sample	# 1	02:16	Subject Test Refused
Breath Volume	0.000	02:16	
Air Blank	0.000	02:17	
Control Test	0.082	02:17	
Air Blank	0.000	02:18	
Diagnostics Check	OK	02:18	

Data Download Date/Time: 02/07/2019 13:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FILBERT,E
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 02/02/2019 05:11:06

Diagnostics Check	OK	05:04
Air Blank	0.000	05:04
Control Test	0.082	05:05
Air Blank	0.000	05:05
Subject Sample # 1	0.113	05:06
Breath Volume	1.449	05:06
Air Blank	0.000	05:07
Air Blank	0.000	05:08
Subject Sample # 2	0.110	05:09
Breath Volume	2.265	05:09
Air Blank	0.000	05:10
Control Test	0.082	05:10
Air Blank	0.000	05:11
Diagnostics Check	OK	05:11

Data Download Date/Time: 02/07/2019 13:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVERETT,I,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 14:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:25

SUBJECT'S LAST NAME: HELD

DATE/TIME OF TEST: 02/02/2019 15:59:41

Diagnostics Check	OK	15:51	
Air Blank	0.000	15:51	
Control Test	0.082	15:51	
Air Blank	0.000	15:52	
Subject Sample	# 1	15:55	No Sample Provided
Breath Volume	0.000	15:55	
Air Blank	0.000	15:56	
Air Blank	0.000	15:58	
Subject Sample	# 2	15:58	Subject Test Refused
Breath Volume	0.000	15:58	
Air Blank	0.000	15:58	
Control Test	0.083	15:59	
Air Blank	0.000	15:59	
Diagnostics Check	OK	15:59	

Data Download Date/Time: 02/07/2019 13:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVERETT,I,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 14:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:25

SUBJECT'S LAST NAME: HELD

DATE/TIME OF TEST: 02/02/2019 16:22:09

Diagnostics Check	OK	16:14
Air Blank	0.000	16:14
Control Test	0.083	16:15
Air Blank	0.000	16:15
Subject Sample # 1	0.133	16:17
Breath Volume	1.605	16:17
Air Blank	0.000	16:17
Air Blank	0.000	16:19
Subject Sample # 2	0.146	16:20
Breath Volume	1.160	16:20
Air Blank	0.000	16:21
Control Test	0.080	16:21
Air Blank	0.000	16:22
Diagnostics Check	OK	16:22

Data Download Date/Time: 02/07/2019 13:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAGE,D,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:03

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 02/03/2019 21:26:57

Diagnostics Check	OK	21:24	
Air Blank	0.000	21:24	
Control Test	0.082	21:24	
Air Blank	0.000	21:25	
Subject Sample	# 1	21:25	Subject Test Refused
Breath Volume	0.000	21:25	
Air Blank	0.000	21:26	
Control Test	0.081	21:26	
Air Blank	0.000	21:26	
Diagnostics Check	OK	21:26	

Data Download Date/Time: 02/07/2019 13:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN, C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: BURT

DATE/TIME OF TEST: 02/04/2019 00:35:27

Diagnostics Check	OK	00:32	
Air Blank	0.000	00:33	
Control Test	0.082	00:33	
Air Blank	0.000	00:33	
Subject Sample	# 1	00:34	Subject Test Refused
Breath Volume	0.000	00:34	
Air Blank	0.000	00:34	
Control Test	0.082	00:34	
Air Blank	0.000	00:35	
Diagnostics Check	OK	00:35	

Data Download Date/Time: 02/07/2019 13:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: Fiset, J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: SOBANJO

DATE/TIME OF TEST: 02/04/2019 01:11:40

Diagnostics Check	OK	01:08	
Air Blank	0.000	01:09	
Control Test	0.083	01:09	
Air Blank	0.000	01:10	
Subject Sample	# 1	01:10	Subject Test Refused
Breath Volume	0.000	01:10	
Air Blank	0.000	01:10	
Control Test	0.082	01:11	
Air Blank	0.000	01:11	
Diagnostics Check	OK	01:11	

Data Download Date/Time: 02/07/2019 13:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CAMACHO,K
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: LECHLEITER

DATE/TIME OF TEST: 02/04/2019 02:02:47

Diagnostics Check	OK	01:59	
Air Blank	0.000	02:00	
Control Test	0.082	02:00	
Air Blank	0.000	02:01	
Subject Sample	# 1	02:01	Subject Test Refused
Breath Volume	0.000	02:01	
Air Blank	0.000	02:01	
Control Test	0.083	02:02	
Air Blank	0.000	02:02	
Diagnostics Check	OK	02:02	

Data Download Date/Time: 02/07/2019 13:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNSTON, L
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: GROVER

DATE/TIME OF TEST: 02/04/2019 02:42:36

Diagnostics Check	OK	02:36
Air Blank	0.000	02:36
Control Test	0.083	02:36
Air Blank	0.000	02:37
Subject Sample # 1	0.164	02:37
Breath Volume	3.027	02:37
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample # 2	0.161	02:40
Breath Volume	3.023	02:40
Air Blank	0.000	02:41
Control Test	0.082	02:41
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 02/07/2019 13:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DIAZ,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: MARTINDALE

DATE/TIME OF TEST: 02/04/2019 04:33:19

Diagnostics Check	OK	04:30	
Air Blank	0.000	04:30	
Control Test	0.082	04:31	
Air Blank	0.000	04:31	
Subject Sample	# 1	04:31	Subject Test Refused
Breath Volume	0.000	04:31	
Air Blank	0.000	04:32	
Control Test	0.082	04:32	
Air Blank	0.000	04:33	
Diagnostics Check	OK	04:33	

Data Download Date/Time: 02/07/2019 13:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ, M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: DASILVA

DATE/TIME OF TEST: 02/05/2019 03:39:26

Diagnostics Check	OK	03:32
Air Blank	0.000	03:32
Control Test	0.083	03:32
Air Blank	0.000	03:33
Subject Sample # 1	0.000	03:34
Breath Volume	1.953	03:34
Air Blank	0.000	03:35
Air Blank	0.000	03:37
Subject Sample # 2	0.000	03:38
Breath Volume	1.535	03:38
Air Blank	0.000	03:38
Control Test	0.082	03:38
Air Blank	0.000	03:39
Diagnostics Check	OK	03:39

Data Download Date/Time: 02/07/2019 13:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VOUMARD,MARCEL,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:18

SUBJECT'S LAST NAME: MOSLEY

DATE/TIME OF TEST: 02/05/2019 21:42:18

Diagnostics Check	OK	21:38	
Air Blank	0.000	21:39	
Control Test	0.080	21:39	
Air Blank	0.000	21:40	
Subject Sample	# 1	21:40	Subject Test Refused
Breath Volume	0.000	21:40	
Air Blank	0.000	21:41	
Control Test	0.081	21:41	
Air Blank	0.000	21:42	
Diagnostics Check	OK	21:42	

Data Download Date/Time: 02/07/2019 13:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEISTHAMMEL, SEAN, P
BREATH TEST OPERATOR AGENCY: COCONUT CREEK PD
ARRESTING OFFICER: BUSTAMANTE, WILLIAM
ARRESTING OFFICER AGENCY: COCONUT CREEK PD
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: DE ANDRADE AMARAL

DATE/TIME OF TEST: 02/04/2019 03:01:44

Diagnostics Check	OK	02:58	
Air Blank	0.000	02:59	
Control Test	0.078	02:59	
Air Blank	0.000	03:00	
Subject Sample	# 1	03:00	Subject Test Refused
Breath Volume	0.000	03:00	
Air Blank	0.000	03:00	
Control Test	0.078	03:01	
Air Blank	0.000	03:01	
Diagnostics Check	OK	03:01	

Data Download Date/Time: 02/27/2019 16:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEISTHAMMEL,SEAN,P
BREATH TEST OPERATOR AGENCY: COCONUT CREEK PD
ARRESTING OFFICER: CALDERON,ANDREA
ARRESTING OFFICER AGENCY: COCONUT CREEK PD
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: WALCOTT

DATE/TIME OF TEST: 02/04/2019 04:01:19

Diagnostics Check	OK	03:58	
Air Blank	0.000	03:58	
Control Test	0.079	03:59	
Air Blank	0.000	03:59	
Subject Sample	# 1	03:59	Subject Test Refused
Breath Volume	0.000	03:59	
Air Blank	0.000	04:00	
Control Test	0.078	04:00	
Air Blank	0.000	04:01	
Diagnostics Check	OK	04:01	

Data Download Date/Time: 02/27/2019 16:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: COCONUT CREEK PD
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: COCONUT CREEK PD
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: KOONS

DATE/TIME OF TEST: 02/09/2019 02:53:48

Diagnostics Check	OK	02:51	
Air Blank	0.000	02:51	
Control Test	0.078	02:51	
Air Blank	0.000	02:52	
Subject Sample	# 1	02:52	Subject Test Refused
Breath Volume	0.000	02:52	
Air Blank	0.000	02:52	
Control Test	0.079	02:53	
Air Blank	0.000	02:53	
Diagnostics Check	OK	02:53	

Data Download Date/Time: 02/27/2019 16:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: COCONUT CREEK PD
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: COCONUT CREEK PD
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: COTO PORTILLO

DATE/TIME OF TEST: 02/10/2019 02:51:13

Diagnostics Check	OK	02:34	
Air Blank	0.000	02:34	
Control Test	0.078	02:35	
Air Blank	0.000	02:35	
Subject Sample	# 1 0.208	02:38	
Breath Volume	1.308	02:38	
Air Blank	0.000	02:38	
Air Blank	0.000	02:40	
Subject Sample	# 2 0.169	02:43	Slope Not Level
Breath Volume	1.328	02:43	
Air Blank	0.000	02:44	
Air Blank	0.000	02:46	
Subject Sample	# 3 0.170	02:49	Slope Not Level
Breath Volume	1.523	02:49	
Air Blank	0.000	02:50	
Control Test	0.077	02:50	
Air Blank	0.000	02:51	
Diagnostics Check	OK	02:51	

Data Download Date/Time: 02/27/2019 16:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: COCONUT CREEK PD
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: COCONUT CREEK PD
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: COTO PORTILLO

DATE/TIME OF TEST: 02/10/2019 02:59:58

Diagnostics Check	OK	02:53
Air Blank	0.000	02:53
Control Test	0.078	02:54
Air Blank	0.000	02:54
Subject Sample # 1	0.186	02:55
Breath Volume	1.925	02:55
Air Blank	0.000	02:56
Air Blank	0.000	02:58
Subject Sample # 2	0.184	02:58
Breath Volume	1.625	02:58
Air Blank	0.000	02:59
Control Test	0.077	02:59
Air Blank	0.000	02:59
Diagnostics Check	OK	02:59

Data Download Date/Time: 02/27/2019 16:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: COCONUT CREEK PD
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: COCONUT CREEK PD
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: AUSTIN

DATE/TIME OF TEST: 02/11/2019 01:33:33

Diagnostics Check	OK	01:27
Air Blank	0.000	01:27
Control Test	0.079	01:27
Air Blank	0.000	01:28
Subject Sample # 1	0.075	01:29
Breath Volume	2.437	01:29
Air Blank	0.000	01:29
Air Blank	0.000	01:31
Subject Sample # 2	0.074	01:32
Breath Volume	3.105	01:32
Air Blank	0.000	01:32
Control Test	0.078	01:32
Air Blank	0.000	01:33
Diagnostics Check	OK	01:33

Data Download Date/Time: 02/27/2019 16:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,DONALD,P
BREATH TEST OPERATOR AGENCY: COCONUT CREEK PD
ARRESTING OFFICER: DAVIS,DONALD
ARRESTING OFFICER AGENCY: COCONUT CREEK PD
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 08:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:08

SUBJECT'S LAST NAME: CARVALHO

DATE/TIME OF TEST: 02/16/2019 08:45:29

Diagnostics Check	OK	08:39
Air Blank	0.000	08:40
Control Test	0.079	08:40
Air Blank	0.000	08:40
Subject Sample # 1	0.000	08:41
Breath Volume	3.332	08:41
Air Blank	0.000	08:41
Air Blank	0.000	08:43
Subject Sample # 2	0.000	08:44
Breath Volume	2.789	08:44
Air Blank	0.000	08:44
Control Test	0.078	08:44
Air Blank	0.000	08:45
Diagnostics Check	OK	08:45

Data Download Date/Time: 02/27/2019 16:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: COCONUT CREEK PD
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: COCONUT CREEK PD
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: ALANKAR

DATE/TIME OF TEST: 02/19/2019 05:08:29

Diagnostics Check	OK	05:05	
Air Blank	0.000	05:05	
Control Test	0.079	05:06	
Air Blank	0.000	05:06	
Subject Sample	# 1	05:07	Subject Test Refused
Breath Volume	0.000	05:07	
Air Blank	0.000	05:07	
Control Test	0.079	05:07	
Air Blank	0.000	05:08	
Diagnostics Check	OK	05:08	

Data Download Date/Time: 02/27/2019 16:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: COCONUT CREEK PD
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: COCONUT CREEK PD
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: MENDEZ ABREU

DATE/TIME OF TEST: 02/23/2019 03:34:31

Diagnostics Check	OK	03:31	
Air Blank	0.000	03:31	
Control Test	0.079	03:32	
Air Blank	0.000	03:32	
Subject Sample	# 1	03:33	Subject Test Refused
Breath Volume	0.000	03:33	
Air Blank	0.000	03:33	
Control Test	0.080	03:33	
Air Blank	0.000	03:34	
Diagnostics Check	OK	03:34	

Data Download Date/Time: 02/27/2019 16:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: COCONUT CREEK PD
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: COCONUT CREEK PD
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 21:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: CANTWELL

DATE/TIME OF TEST: 02/23/2019 23:37:24

Diagnostics Check	OK	23:34	
Air Blank	0.000	23:35	
Control Test	0.079	23:35	
Air Blank	0.000	23:35	
Subject Sample	# 1	23:36	Subject Test Refused
Breath Volume	0.000	23:36	
Air Blank	0.000	23:36	
Control Test	0.079	23:36	
Air Blank	0.000	23:37	
Diagnostics Check	OK	23:37	

Data Download Date/Time: 02/27/2019 16:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: COCONUT CREEK PD
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: COCONUT CREEK PD
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: VALDEZ ROMAN

DATE/TIME OF TEST: 02/24/2019 22:24:16

Diagnostics Check	OK	22:21	
Air Blank	0.000	22:21	
Control Test	0.077	22:22	
Air Blank	0.000	22:22	
Subject Sample	# 1	22:22	Subject Test Refused
Breath Volume	0.000	22:22	
Air Blank	0.000	22:23	
Control Test	0.079	22:23	
Air Blank	0.000	22:24	
Diagnostics Check	OK	22:24	

Data Download Date/Time: 02/27/2019 16:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAY,L,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KARPINSKI,C
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:54

SUBJECT'S LAST NAME: RIELLY

DATE/TIME OF TEST: 02/09/2019 22:24:36

Diagnostics Check	OK	22:14
Air Blank	0.000	22:15
Control Test	0.080	22:15
Air Blank	0.000	22:16
Subject Sample # 1	0.193	22:16
Breath Volume	2.191	22:16
Air Blank	0.000	22:17
Air Blank	0.000	22:19
Subject Sample # 2	0.164	22:20
Breath Volume	1.648	22:20
Air Blank	0.000	22:20
Air Blank	0.000	22:22
Subject Sample # 3	0.184	22:22
Breath Volume	2.781	22:22
Air Blank	0.000	22:23
Control Test	0.078	22:23
Air Blank	0.000	22:24
Diagnostics Check	OK	22:24

Data Download Date/Time: 03/04/2019 11:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVERETT,I,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: PRITCHARD

DATE/TIME OF TEST: 02/10/2019 21:18:05

Diagnostics Check	OK	21:15	
Air Blank	0.000	21:15	
Control Test	0.079	21:16	
Air Blank	0.000	21:16	
Subject Sample	# 1	21:16	Subject Test Refused
Breath Volume	0.000	21:16	
Air Blank	0.000	21:17	
Control Test	0.079	21:17	
Air Blank	0.000	21:18	
Diagnostics Check	OK	21:18	

Data Download Date/Time: 03/04/2019 11:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVERETT,I,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: VANN

DATE/TIME OF TEST: 02/10/2019 22:10:21

Diagnostics Check	OK	21:58	
Air Blank	0.000	21:58	
Control Test	0.080	21:59	
Air Blank	0.000	21:59	
Subject Sample	# 1 0.131	22:00	
Breath Volume	1.312	22:00	
Air Blank	0.000	22:00	
Air Blank	0.000	22:02	
Subject Sample	# 2 0.102	22:05	Slope Not Level
Breath Volume	1.171	22:05	
Air Blank	0.000	22:06	
Air Blank	0.000	22:08	
Subject Sample	# 3	22:08	Subject Test Refused
Breath Volume	0.089	22:08	
Air Blank	0.000	22:09	
Control Test	0.076	22:09	
Air Blank	0.000	22:10	
Diagnostics Check	OK	22:10	

Data Download Date/Time: 03/04/2019 11:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AUGUSTINE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HYSMITH,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 02/11/2019 04:42:54

Diagnostics Check	OK	04:40	
Air Blank	0.000	04:40	
Control Test	0.080	04:40	
Air Blank	0.000	04:41	
Subject Sample	# 1	04:41	Subject Test Refused
Breath Volume	0.000	04:41	
Air Blank	0.000	04:41	
Control Test	0.079	04:42	
Air Blank	0.000	04:42	
Diagnostics Check	OK	04:42	

Data Download Date/Time: 03/04/2019 11:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ, M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: ARGROVES

DATE/TIME OF TEST: 02/13/2019 00:39:36

Diagnostics Check	OK	00:32
Air Blank	0.000	00:33
Control Test	0.079	00:33
Air Blank	0.000	00:33
Subject Sample # 1	0.120	00:34
Breath Volume	2.398	00:34
Air Blank	0.000	00:35
Air Blank	0.000	00:37
Subject Sample # 2	0.118	00:38
Breath Volume	2.320	00:38
Air Blank	0.000	00:38
Control Test	0.079	00:38
Air Blank	0.000	00:39
Diagnostics Check	OK	00:39

Data Download Date/Time: 03/04/2019 11:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: BLEAKLEY

DATE/TIME OF TEST: 02/13/2019 03:00:04

Diagnostics Check	OK	02:57	
Air Blank	0.000	02:57	
Control Test	0.079	02:57	
Air Blank	0.000	02:58	
Subject Sample	# 1	02:58	Subject Test Refused
Breath Volume	0.000	02:58	
Air Blank	0.000	02:59	
Control Test	0.079	02:59	
Air Blank	0.000	02:59	
Diagnostics Check	OK	03:00	

Data Download Date/Time: 03/04/2019 11:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAY,L,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:49

SUBJECT'S LAST NAME: TEESATESKI

DATE/TIME OF TEST: 02/16/2019 18:13:56

Diagnostics Check	OK	18:10	
Air Blank	0.000	18:10	
Control Test	0.080	18:10	
Air Blank	0.000	18:11	
Subject Sample	# 1	18:12	Subject Test Refused
Breath Volume	0.000	18:12	
Air Blank	0.000	18:13	
Control Test	0.079	18:13	
Air Blank	0.000	18:13	
Diagnostics Check	OK	18:13	

Data Download Date/Time: 03/04/2019 11:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAY,L,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRYANT,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: OSTEN

DATE/TIME OF TEST: 02/16/2019 22:34:04

Diagnostics Check	OK	22:31	
Air Blank	0.000	22:31	
Control Test	0.080	22:31	
Air Blank	0.000	22:32	
Subject Sample	# 1	22:32	Subject Test Refused
Breath Volume	0.000	22:32	
Air Blank	0.000	22:33	
Control Test	0.079	22:33	
Air Blank	0.000	22:33	
Diagnostics Check	OK	22:34	

Data Download Date/Time: 03/04/2019 11:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAY,L,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 02/17/2019 20:55:26

Diagnostics Check	OK	20:49
Air Blank	0.000	20:49
Control Test	0.080	20:50
Air Blank	0.000	20:50
Subject Sample # 1	0.000	20:51
Breath Volume	2.066	20:51
Air Blank	0.000	20:51
Air Blank	0.000	20:53
Subject Sample # 2	0.000	20:54
Breath Volume	1.417	20:54
Air Blank	0.000	20:54
Control Test	0.079	20:54
Air Blank	0.000	20:55
Diagnostics Check	OK	20:55

Data Download Date/Time: 03/04/2019 11:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAY,L,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: SHAQAREQ

DATE/TIME OF TEST: 02/17/2019 21:38:56

Diagnostics Check	OK	21:36	
Air Blank	0.000	21:36	
Control Test	0.079	21:36	
Air Blank	0.000	21:37	
Subject Sample	# 1	21:37	Subject Test Refused
Breath Volume	0.000	21:37	
Air Blank	0.000	21:38	
Control Test	0.079	21:38	
Air Blank	0.000	21:38	
Diagnostics Check	OK	21:38	

Data Download Date/Time: 03/04/2019 11:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVERETT,I,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: STRICKLAND

DATE/TIME OF TEST: 02/19/2019 18:13:54

Diagnostics Check	OK	18:07
Air Blank	0.000	18:07
Control Test	0.080	18:08
Air Blank	0.000	18:08
Subject Sample # 1	0.241	18:09
Breath Volume	2.410	18:09
Air Blank	0.000	18:09
Air Blank	0.000	18:11
Subject Sample # 2	0.238	18:12
Breath Volume	2.558	18:12
Air Blank	0.000	18:12
Control Test	0.079	18:13
Air Blank	0.000	18:13
Diagnostics Check	OK	18:13

Data Download Date/Time: 03/04/2019 11:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVERETT,I,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRYANT,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: LAMBERT

DATE/TIME OF TEST: 02/19/2019 23:02:11

Diagnostics Check	OK	22:55
Air Blank	0.000	22:56
Control Test	0.080	22:56
Air Blank	0.000	22:56
Subject Sample # 1	0.000	22:57
Breath Volume	1.824	22:57
Air Blank	0.000	22:58
Air Blank	0.000	23:00
Subject Sample # 2	0.000	23:00
Breath Volume	1.292	23:00
Air Blank	0.000	23:01
Control Test	0.080	23:01
Air Blank	0.000	23:02
Diagnostics Check	OK	23:02

Data Download Date/Time: 03/04/2019 11:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVERETT,I,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: LUMPKIN

DATE/TIME OF TEST: 02/20/2019 22:42:26

Diagnostics Check	OK	22:35
Air Blank	0.000	22:36
Control Test	0.079	22:36
Air Blank	0.000	22:37
Subject Sample # 1	0.000	22:37
Breath Volume	1.843	22:37
Air Blank	0.000	22:38
Air Blank	0.000	22:40
Subject Sample # 2	0.000	22:41
Breath Volume	1.894	22:41
Air Blank	0.000	22:41
Control Test	0.079	22:41
Air Blank	0.000	22:42
Diagnostics Check	OK	22:42

Data Download Date/Time: 03/04/2019 11:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DIAZ,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 02/22/2019 00:37:38

Diagnostics Check	OK	00:34	
Air Blank	0.000	00:35	
Control Test	0.080	00:35	
Air Blank	0.000	00:35	
Subject Sample	# 1	00:36	Subject Test Refused
Breath Volume	0.000	00:36	
Air Blank	0.000	00:36	
Control Test	0.079	00:37	
Air Blank	0.000	00:37	
Diagnostics Check	OK	00:37	

Data Download Date/Time: 03/04/2019 11:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BASS,HUNTER,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: HICKEY

DATE/TIME OF TEST: 02/22/2019 02:20:28

Diagnostics Check	OK	02:13
Air Blank	0.000	02:14
Control Test	0.079	02:14
Air Blank	0.000	02:15
Subject Sample # 1	0.197	02:15
Breath Volume	4.097	02:15
Air Blank	0.000	02:16
Air Blank	0.000	02:18
Subject Sample # 2	0.192	02:18
Breath Volume	2.140	02:18
Air Blank	0.000	02:19
Control Test	0.078	02:19
Air Blank	0.000	02:20
Diagnostics Check	OK	02:20

Data Download Date/Time: 03/04/2019 11:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GIBSON,D
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:49

SUBJECT'S LAST NAME: HAGEN

DATE/TIME OF TEST: 02/22/2019 05:14:28

Diagnostics Check	OK	05:11	
Air Blank	0.000	05:12	
Control Test	0.079	05:12	
Air Blank	0.000	05:12	
Subject Sample	# 1	05:13	Subject Test Refused
Breath Volume	0.000	05:13	
Air Blank	0.000	05:13	
Control Test	0.079	05:13	
Air Blank	0.000	05:14	
Diagnostics Check	OK	05:14	

Data Download Date/Time: 03/04/2019 11:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVERETT,I,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: ROBERTSON

DATE/TIME OF TEST: 02/22/2019 22:11:48

Diagnostics Check	OK	22:05
Air Blank	0.000	22:05
Control Test	0.079	22:05
Air Blank	0.000	22:06
Subject Sample # 1	0.107	22:07
Breath Volume	2.933	22:07
Air Blank	0.000	22:07
Air Blank	0.000	22:09
Subject Sample # 2	0.107	22:10
Breath Volume	2.210	22:10
Air Blank	0.000	22:10
Control Test	0.079	22:11
Air Blank	0.000	22:11
Diagnostics Check	OK	22:11

Data Download Date/Time: 03/04/2019 11:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISSETTE, J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 02/23/2019 01:30:46

Diagnostics Check	OK	01:27	
Air Blank	0.000	01:28	
Control Test	0.080	01:28	
Air Blank	0.000	01:29	
Subject Sample	# 1	01:29	Subject Test Refused
Breath Volume	0.000	01:29	
Air Blank	0.000	01:29	
Control Test	0.079	01:30	
Air Blank	0.000	01:30	
Diagnostics Check	OK	01:30	

Data Download Date/Time: 03/04/2019 11:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 02/23/2019 04:57:08

Diagnostics Check	OK	04:49
Air Blank	0.000	04:50
Control Test	0.079	04:50
Air Blank	0.000	04:50
Subject Sample # 1	0.131	04:51
Breath Volume	1.210	04:51
Air Blank	0.000	04:52
Air Blank	0.000	04:54
Subject Sample # 2	0.130	04:55
Breath Volume	1.464	04:55
Air Blank	0.000	04:56
Control Test	0.079	04:56
Air Blank	0.000	04:57
Diagnostics Check	OK	04:57

Data Download Date/Time: 03/04/2019 11:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PATTERSON, K
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:59

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 02/23/2019 05:28:32

Diagnostics Check	OK	05:20	
Air Blank	0.000	05:20	
Control Test	0.079	05:20	
Air Blank	0.000	05:21	
Subject Sample	# 1	05:24	No Sample Provided
Breath Volume	0.000	05:24	
Air Blank	0.000	05:24	
Air Blank	0.000	05:27	
Subject Sample	# 2	05:27	Subject Test Refused
Breath Volume	0.000	05:27	
Air Blank	0.000	05:27	
Control Test	0.079	05:27	
Air Blank	0.000	05:28	
Diagnostics Check	OK	05:28	

Data Download Date/Time: 03/04/2019 11:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VOUMARD,MARCEL,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DIAZ,C.A.
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:07

SUBJECT'S LAST NAME: VENKATASUBBAN

DATE/TIME OF TEST: 02/23/2019 21:35:09

Diagnostics Check	OK	21:27
Air Blank	0.000	21:28
Control Test	0.080	21:28
Air Blank	0.000	21:29
Subject Sample # 1	0.180	21:30
Breath Volume	2.066	21:30
Air Blank	0.000	21:31
Air Blank	0.000	21:33
Subject Sample # 2	0.183	21:33
Breath Volume	1.960	21:33
Air Blank	0.000	21:34
Control Test	0.078	21:34
Air Blank	0.000	21:35
Diagnostics Check	OK	21:35

Data Download Date/Time: 03/04/2019 11:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY, SHANON, E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JARMAN, S
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: SIWEK

DATE/TIME OF TEST: 02/24/2019 04:40:16

Diagnostics Check	OK	04:25	
Air Blank	0.000	04:25	
Control Test	0.079	04:26	
Air Blank	0.000	04:26	
Subject Sample	# 1 0.133	04:27	
Breath Volume	1.246	04:27	
Air Blank	0.000	04:28	
Air Blank	0.000	04:29	
Subject Sample	# 2 0.121	04:32	Volume Not Met
Breath Volume	0.125	04:32	
Air Blank	0.000	04:33	
Air Blank	0.000	04:35	
Subject Sample	# 3 0.096	04:38	Volume Not Met
Breath Volume	0.234	04:38	
Air Blank	0.000	04:39	
Control Test	0.077	04:39	
Air Blank	0.000	04:40	
Diagnostics Check	OK	04:40	

Data Download Date/Time: 03/04/2019 11:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVERETT,I,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: HOLLIMAN

DATE/TIME OF TEST: 02/24/2019 21:47:05

Diagnostics Check	OK	21:41
Air Blank	0.000	21:41
Control Test	0.079	21:41
Air Blank	0.000	21:42
Subject Sample # 1	0.000	21:42
Breath Volume	2.070	21:42
Air Blank	0.000	21:43
Air Blank	0.000	21:45
Subject Sample # 2	0.000	21:45
Breath Volume	1.980	21:45
Air Blank	0.000	21:46
Control Test	0.080	21:46
Air Blank	0.000	21:46
Diagnostics Check	OK	21:47

Data Download Date/Time: 03/04/2019 11:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEURER,DENNIS,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: BROADWAY

DATE/TIME OF TEST: 02/24/2019 23:02:54

Diagnostics Check	OK	22:55
Air Blank	0.000	22:55
Control Test	0.079	22:56
Air Blank	0.000	22:56
Subject Sample # 1	0.155	22:58
Breath Volume	1.613	22:58
Air Blank	0.000	22:58
Air Blank	0.000	23:00
Subject Sample # 2	0.159	23:01
Breath Volume	1.597	23:01
Air Blank	0.000	23:01
Control Test	0.078	23:02
Air Blank	0.000	23:02
Diagnostics Check	OK	23:02

Data Download Date/Time: 03/04/2019 11:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VOUMARD,MARCEL,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J.M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: CONROY

DATE/TIME OF TEST: 02/25/2019 22:37:22

Diagnostics Check	OK	22:30
Air Blank	0.000	22:31
Control Test	0.080	22:31
Air Blank	0.000	22:32
Subject Sample # 1	0.075	22:32
Breath Volume	5.605	22:32
Air Blank	0.000	22:33
Air Blank	0.000	22:35
Subject Sample # 2	0.060	22:35
Breath Volume	1.296	22:35
Air Blank	0.000	22:36
Control Test	0.079	22:36
Air Blank	0.000	22:37
Diagnostics Check	OK	22:37

Data Download Date/Time: 03/04/2019 11:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: MIXON

DATE/TIME OF TEST: 02/26/2019 00:33:26

Diagnostics Check	OK	00:23
Air Blank	0.000	00:24
Control Test	0.079	00:24
Air Blank	0.000	00:24
Subject Sample # 1	0.180	00:25
Breath Volume	3.320	00:25
Air Blank	0.000	00:26
Air Blank	0.000	00:28
Subject Sample # 2	0.156	00:28
Breath Volume	1.242	00:28
Air Blank	0.000	00:29
Air Blank	0.000	00:31
Subject Sample # 3	0.176	00:31
Breath Volume	2.367	00:31
Air Blank	0.000	00:32
Control Test	0.079	00:32
Air Blank	0.000	00:33
Diagnostics Check	OK	00:33

Data Download Date/Time: 03/04/2019 11:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE, J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: PARIS

DATE/TIME OF TEST: 02/26/2019 03:06:33

Diagnostics Check	OK	03:00
Air Blank	0.000	03:00
Control Test	0.079	03:00
Air Blank	0.000	03:01
Subject Sample # 1	0.000	03:02
Breath Volume	2.195	03:02
Air Blank	0.000	03:02
Air Blank	0.000	03:04
Subject Sample # 2	0.000	03:05
Breath Volume	2.128	03:05
Air Blank	0.000	03:05
Control Test	0.079	03:05
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

Data Download Date/Time: 03/04/2019 11:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVERETT,I,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:16

SUBJECT'S LAST NAME: BUKOV

DATE/TIME OF TEST: 02/28/2019 19:44:03

Diagnostics Check	OK	19:37
Air Blank	0.000	19:37
Control Test	0.079	19:38
Air Blank	0.000	19:38
Subject Sample # 1	0.079	19:39
Breath Volume	1.500	19:39
Air Blank	0.000	19:39
Air Blank	0.000	19:41
Subject Sample # 2	0.081	19:42
Breath Volume	1.859	19:42
Air Blank	0.000	19:43
Control Test	0.079	19:43
Air Blank	0.000	19:43
Diagnostics Check	OK	19:44

Data Download Date/Time: 03/04/2019 11:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASLINE,T'NEISHA,S
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: HART,B
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 01/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: HAMPTON

DATE/TIME OF TEST: 02/02/2019 02:03:50

Diagnostics Check	OK	01:52	
Air Blank	0.000	01:52	
Control Test	0.080	01:53	
Air Blank	0.000	01:53	
Subject Sample	# 1	01:56	No Sample Provided
Breath Volume	0.000	01:56	
Air Blank	0.000	01:57	
Air Blank	0.000	01:59	
Subject Sample	# 2	02:02	No Sample Provided
Breath Volume	0.000	02:02	
Air Blank	0.000	02:02	
Control Test	0.081	02:03	
Air Blank	0.000	02:03	
Diagnostics Check	OK	02:03	

Data Download Date/Time: 02/22/2019 10:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASLINE,T'NEISHA,S
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: LANIER,H
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 01/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 02/02/2019 21:31:30

Diagnostics Check	OK	21:18	
Air Blank	0.000	21:18	
Control Test	0.079	21:19	
Air Blank	0.000	21:19	
Subject Sample	# 1 0.242	21:20	No .020 agreement
Breath Volume	1.160	21:20	
Air Blank	0.000	21:21	
Air Blank	0.000	21:23	
Subject Sample	# 2 0.227	21:26	Slope Not Level
Breath Volume	1.285	21:26	
Air Blank	0.000	21:27	
Air Blank	0.000	21:29	
Subject Sample	# 3 0.279	21:29	No .020 agreement
Breath Volume	1.421	21:29	
Air Blank	0.000	21:30	
Control Test	0.078	21:30	
Air Blank	0.000	21:31	
Diagnostics Check	OK	21:31	

Data Download Date/Time: 02/22/2019 10:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTORO,C,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: MULBERRY IV

DATE/TIME OF TEST: 02/03/2019 01:12:28

Diagnostics Check	OK	01:09	
Air Blank	0.000	01:10	
Control Test	0.080	01:10	
Air Blank	0.000	01:10	
Subject Sample	# 1	01:11	Subject Test Refused
Breath Volume	0.000	01:11	
Air Blank	0.000	01:11	
Control Test	0.079	01:11	
Air Blank	0.000	01:12	
Diagnostics Check	OK	01:12	

Data Download Date/Time: 02/22/2019 10:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTORO,C,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: DAWSON

DATE/TIME OF TEST: 02/03/2019 04:21:23

Diagnostics Check	OK	04:13
Air Blank	0.000	04:13
Control Test	0.079	04:13
Air Blank	0.000	04:14
Subject Sample # 1	0.185	04:15
Breath Volume	2.042	04:15
Air Blank	0.000	04:15
Air Blank	0.000	04:17
Subject Sample # 2	0.172	04:19
Breath Volume	1.347	04:19
Air Blank	0.000	04:20
Control Test	0.079	04:20
Air Blank	0.000	04:21
Diagnostics Check	OK	04:21

Data Download Date/Time: 02/22/2019 10:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,C,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: DRAGOMIR,I
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: THEROUX

DATE/TIME OF TEST: 02/04/2019 00:54:09

Diagnostics Check	OK	00:46
Air Blank	0.000	00:47
Control Test	0.080	00:47
Air Blank	0.000	00:48
Subject Sample # 1	0.167	00:49
Breath Volume	2.691	00:49
Air Blank	0.000	00:50
Air Blank	0.000	00:52
Subject Sample # 2	0.169	00:52
Breath Volume	1.878	00:52
Air Blank	0.000	00:53
Control Test	0.079	00:53
Air Blank	0.000	00:54
Diagnostics Check	OK	00:54

Data Download Date/Time: 02/22/2019 10:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARLOW,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: PRITCHARD,TRAVIS
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:22

SUBJECT'S LAST NAME: KENWORTHY

DATE/TIME OF TEST: 02/04/2019 19:56:43

Diagnostics Check	OK	19:47	
Air Blank	0.000	19:47	
Control Test	0.080	19:47	
Air Blank	0.000	19:48	
Subject Sample	# 1	19:51	No Sample Provided
Breath Volume	0.000	19:51	
Air Blank	0.000	19:51	
Air Blank	0.000	19:53	
Subject Sample	# 2	19:55	Subject Test Refused
Breath Volume	0.000	19:55	
Air Blank	0.000	19:55	
Control Test	0.079	19:56	
Air Blank	0.000	19:56	
Diagnostics Check	OK	19:56	

Data Download Date/Time: 02/22/2019 10:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARLOW,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: FALL,JORDAN
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: HYLER

DATE/TIME OF TEST: 02/06/2019 00:54:06

Diagnostics Check	OK	00:47
Air Blank	0.000	00:47
Control Test	0.080	00:48
Air Blank	0.000	00:48
Subject Sample # 1	0.220	00:49
Breath Volume	2.726	00:49
Air Blank	0.000	00:50
Air Blank	0.000	00:51
Subject Sample # 2	0.225	00:52
Breath Volume	2.714	00:52
Air Blank	0.000	00:53
Control Test	0.079	00:53
Air Blank	0.000	00:54
Diagnostics Check	OK	00:54

Data Download Date/Time: 02/22/2019 10:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASLINE,T'NEISHA,S
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: HART,B
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 01/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 02/07/2019 03:06:54

Diagnostics Check	OK	02:56
Air Blank	0.000	02:56
Control Test	0.079	02:56
Air Blank	0.000	02:57
Subject Sample # 1	0.229	02:59
Breath Volume	1.500	02:59
Air Blank	0.000	03:00
Air Blank	0.000	03:02
Subject Sample # 2	0.219	03:05
Breath Volume	1.324	03:05
Air Blank	0.000	03:05
Control Test	0.079	03:06
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

Data Download Date/Time: 02/22/2019 10:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARLOW,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GRACKI,NICHOLAS
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: CERCY

DATE/TIME OF TEST: 02/09/2019 04:32:46

Diagnostics Check	OK	04:25
Air Blank	0.000	04:26
Control Test	0.080	04:26
Air Blank	0.000	04:27
Subject Sample # 1	0.086	04:28
Breath Volume	2.886	04:28
Air Blank	0.000	04:28
Air Blank	0.000	04:30
Subject Sample # 2	0.087	04:31
Breath Volume	3.066	04:31
Air Blank	0.000	04:31
Control Test	0.078	04:32
Air Blank	0.000	04:32
Diagnostics Check	OK	04:32

Data Download Date/Time: 02/22/2019 10:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENTZ,JARROD,C
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: CLOUSER,B
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:03

SUBJECT'S LAST NAME: CURRAN

DATE/TIME OF TEST: 02/09/2019 21:34:03

Diagnostics Check	OK	21:26
Air Blank	0.000	21:27
Control Test	0.080	21:27
Air Blank	0.000	21:28
Subject Sample # 1	0.219	21:28
Breath Volume	1.992	21:28
Air Blank	0.000	21:29
Air Blank	0.000	21:30
Subject Sample # 2	0.212	21:32
Breath Volume	1.398	21:32
Air Blank	0.000	21:33
Control Test	0.079	21:33
Air Blank	0.000	21:33
Diagnostics Check	OK	21:34

Data Download Date/Time: 02/22/2019 10:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTORO,CAMERON,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:54

SUBJECT'S LAST NAME: SAWDO

DATE/TIME OF TEST: 02/11/2019 22:27:44

Diagnostics Check	OK	22:21
Air Blank	0.000	22:21
Control Test	0.080	22:22
Air Blank	0.000	22:22
Subject Sample # 1	0.096	22:23
Breath Volume	2.085	22:23
Air Blank	0.000	22:23
Air Blank	0.000	22:25
Subject Sample # 2	0.098	22:26
Breath Volume	1.707	22:26
Air Blank	0.000	22:26
Control Test	0.079	22:27
Air Blank	0.000	22:27
Diagnostics Check	OK	22:27

Data Download Date/Time: 02/22/2019 10:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENTZ,JARROD,C
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: CRUZ,C
ARRESTING OFFICER AGENCY: Green Cove Springs
LAST AGENCY INSPECTION: 01/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: TOMA

DATE/TIME OF TEST: 02/14/2019 03:11:23

Diagnostics Check	OK	03:07	
Air Blank	0.000	03:08	
Control Test	0.080	03:08	
Air Blank	0.000	03:08	
Subject Sample	# 1	03:10	Subject Test Refused
Breath Volume	0.000	03:10	
Air Blank	0.000	03:10	
Control Test	0.080	03:10	
Air Blank	0.000	03:11	
Diagnostics Check	OK	03:11	

Data Download Date/Time: 02/22/2019 10:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNOW,C,D
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GANN,R
ARRESTING OFFICER AGENCY: Green Cove Springs
LAST AGENCY INSPECTION: 01/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 10:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:46

SUBJECT'S LAST NAME: HERSHKOPF

DATE/TIME OF TEST: 02/14/2019 12:19:31

Diagnostics Check	OK	12:13
Air Blank	0.000	12:13
Control Test	0.080	12:14
Air Blank	0.000	12:14
Subject Sample # 1	0.000	12:15
Breath Volume	3.523	12:15
Air Blank	0.000	12:15
Air Blank	0.000	12:17
Subject Sample # 2	0.000	12:18
Breath Volume	3.003	12:18
Air Blank	0.000	12:18
Control Test	0.080	12:18
Air Blank	0.000	12:19
Diagnostics Check	OK	12:19

Data Download Date/Time: 02/22/2019 10:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTORO,CAMERON,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:11

SUBJECT'S LAST NAME: LABAW

DATE/TIME OF TEST: 02/15/2019 20:44:16

Diagnostics Check	OK	20:37
Air Blank	0.000	20:37
Control Test	0.080	20:37
Air Blank	0.000	20:38
Subject Sample # 1	0.124	20:39
Breath Volume	1.890	20:39
Air Blank	0.000	20:40
Air Blank	0.000	20:42
Subject Sample # 2	0.123	20:42
Breath Volume	1.863	20:42
Air Blank	0.000	20:43
Control Test	0.079	20:43
Air Blank	0.000	20:44
Diagnostics Check	OK	20:44

Data Download Date/Time: 02/22/2019 10:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,C,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: SPAULDING,J
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: CRAVEN

DATE/TIME OF TEST: 02/15/2019 23:03:47

Diagnostics Check	OK	22:54	
Air Blank	0.000	22:55	
Control Test	0.080	22:55	
Air Blank	0.000	22:56	
Subject Sample	# 1	22:59	No Sample Provided
Breath Volume	0.000	22:59	
Air Blank	0.000	22:59	
Air Blank	0.000	23:01	
Subject Sample	# 2	23:02	Subject Test Refused
Breath Volume	0.000	23:02	
Air Blank	0.000	23:02	
Control Test	0.080	23:03	
Air Blank	0.000	23:03	
Diagnostics Check	OK	23:03	

Data Download Date/Time: 02/22/2019 10:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASLINE,T'NEISHA,S
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: BUSBEE

DATE/TIME OF TEST: 02/16/2019 00:46:09

Diagnostics Check	OK	00:39
Air Blank	0.000	00:40
Control Test	0.080	00:40
Air Blank	0.000	00:41
Subject Sample # 1	0.220	00:41
Breath Volume	2.261	00:41
Air Blank	0.000	00:42
Air Blank	0.000	00:44
Subject Sample # 2	0.218	00:44
Breath Volume	2.085	00:44
Air Blank	0.000	00:45
Control Test	0.080	00:45
Air Blank	0.000	00:46
Diagnostics Check	OK	00:46

Data Download Date/Time: 02/22/2019 10:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTSTATT,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: TIMS

DATE/TIME OF TEST: 02/16/2019 04:02:30

Diagnostics Check	OK	03:59	
Air Blank	0.000	03:59	
Control Test	0.080	04:00	
Air Blank	0.000	04:00	
Subject Sample	# 1	04:01	Subject Test Refused
Breath Volume	0.000	04:01	
Air Blank	0.000	04:01	
Control Test	0.079	04:01	
Air Blank	0.000	04:02	
Diagnostics Check	OK	04:02	

Data Download Date/Time: 02/22/2019 10:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTSTATT,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: SMITH,T
ARRESTING OFFICER AGENCY: Green Cove Springs
LAST AGENCY INSPECTION: 01/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:24

SUBJECT'S LAST NAME: PIERCE

DATE/TIME OF TEST: 02/21/2019 21:54:00

Diagnostics Check	OK	21:45	
Air Blank	0.000	21:45	
Control Test	0.079	21:46	
Air Blank	0.000	21:46	
Subject Sample	# 1	21:49	No Sample Provided
Breath Volume	0.000	21:49	
Air Blank	0.000	21:50	
Air Blank	0.000	21:52	
Subject Sample	# 2	21:52	Subject Test Refused
Breath Volume	0.000	21:52	
Air Blank	0.000	21:53	
Control Test	0.079	21:53	
Air Blank	0.000	21:53	
Diagnostics Check	OK	21:54	

Data Download Date/Time: 02/22/2019 10:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENTZ,JARROD,C
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: BRICKEY,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: BURROUGHS

DATE/TIME OF TEST: 02/22/2019 21:53:07

Diagnostics Check	OK	21:41	
Air Blank	0.000	21:41	
Control Test	0.079	21:42	
Air Blank	0.000	21:42	
Subject Sample	# 1 0.150	21:43	
Breath Volume	1.234	21:43	
Air Blank	0.000	21:44	
Air Blank	0.000	21:45	
Subject Sample	# 2	21:49	No Sample Provided
Breath Volume	0.000	21:49	
Air Blank	0.000	21:49	
Air Blank	0.000	21:51	
Subject Sample	# 3	21:51	Subject Test Refused
Breath Volume	0.000	21:51	
Air Blank	0.000	21:52	
Control Test	0.080	21:52	
Air Blank	0.000	21:53	
Diagnostics Check	OK	21:53	

Data Download Date/Time: 03/25/2019 12:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENTZ,JARROD,C
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: PRITCHARD,T
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 02/28/2019 23:11:01

Diagnostics Check	OK	23:10	
Air Blank		23:10	RFI Detect
Air Blank	0.000	23:11	

Data Download Date/Time: 03/25/2019 12:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH-LONG,JEFFERSON,D
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: SMITH-LONG,JEFFERSON
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: VILLANUEVA

DATE/TIME OF TEST: 02/16/2019 03:01:40

Diagnostics Check	OK	02:55
Air Blank	0.000	02:55
Control Test	0.080	02:56
Air Blank	0.000	02:56
Subject Sample # 1	0.231	02:57
Breath Volume	2.582	02:57
Air Blank	0.000	02:57
Air Blank	0.000	02:59
Subject Sample # 2	0.227	03:00
Breath Volume	2.109	03:00
Air Blank	0.000	03:00
Control Test	0.079	03:01
Air Blank	0.000	03:01
Diagnostics Check	OK	03:01

Data Download Date/Time: 02/19/2019 16:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVARADO,ANGEL,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ALVARADO,ANGEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: CHANCEY

DATE/TIME OF TEST: 02/23/2019 16:56:42

Diagnostics Check	OK	16:53	
Air Blank	0.000	16:53	
Control Test	0.080	16:54	
Air Blank	0.000	16:54	
Subject Sample	# 1	16:55	Subject Test Refused
Breath Volume	0.000	16:55	
Air Blank	0.000	16:55	
Control Test	0.080	16:56	
Air Blank	0.000	16:56	
Diagnostics Check	OK	16:56	

Data Download Date/Time: 03/31/2019 15:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,TAYLOR,C
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: ANDERSON,TAYLOR
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 01/31/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: RAMNARINE

DATE/TIME OF TEST: 02/17/2019 19:44:46

Diagnostics Check	OK	19:38
Air Blank	0.000	19:38
Control Test	0.077	19:38
Air Blank	0.000	19:39
Subject Sample # 1	0.303	19:40
Breath Volume	2.902	19:40
Air Blank	0.000	19:40
Air Blank	0.000	19:42
Subject Sample # 2	0.287	19:43
Breath Volume	2.007	19:43
Air Blank	0.000	19:43
Control Test	0.076	19:44
Air Blank	0.000	19:44
Diagnostics Check	OK	19:44

Data Download Date/Time: 02/28/2019 15:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARA-VALDERRAMA,MARIO,A
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: VARA-VALDERRAMA,MARIO
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 01/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 02/18/2019 22:48:12

Diagnostics Check	OK	22:41
Air Blank	0.000	22:42
Control Test	0.078	22:42
Air Blank	0.000	22:42
Subject Sample # 1	0.105	22:43
Breath Volume	2.792	22:43
Air Blank	0.000	22:44
Air Blank	0.000	22:46
Subject Sample # 2	0.105	22:46
Breath Volume	3.363	22:46
Air Blank	0.000	22:47
Control Test	0.077	22:47
Air Blank	0.000	22:48
Diagnostics Check	OK	22:48

Data Download Date/Time: 02/28/2019 15:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARA-VALDERRAMA,MARIO,A
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: ZAPATA,RINA
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 01/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: DE JESUS JR

DATE/TIME OF TEST: 02/25/2019 00:11:24

Diagnostics Check	OK	00:02
Air Blank	0.000	00:02
Control Test	0.076	00:02
Air Blank	0.000	00:03
Subject Sample # 1	0.273	00:04
Breath Volume	1.222	00:04
Air Blank	0.000	00:05
Air Blank	0.000	00:07
Subject Sample # 2	0.260	00:09
Breath Volume	1.925	00:09
Air Blank	0.000	00:10
Control Test	0.075	00:10
Air Blank	0.000	00:11
Diagnostics Check	OK	00:11

Data Download Date/Time: 02/28/2019 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURDOCK,TROY,J
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: HOLZ,CHRISTOPHER
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 01/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: MCKENZIE

DATE/TIME OF TEST: 02/27/2019 03:55:33

Diagnostics Check	OK	03:49
Air Blank	0.000	03:49
Control Test	0.078	03:49
Air Blank	0.000	03:50
Subject Sample # 1	0.134	03:51
Breath Volume	2.902	03:51
Air Blank	0.000	03:51
Air Blank	0.000	03:53
Subject Sample # 2	0.133	03:54
Breath Volume	2.015	03:54
Air Blank	0.000	03:54
Control Test	0.077	03:54
Air Blank	0.000	03:55
Diagnostics Check	OK	03:55

Data Download Date/Time: 02/28/2019 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LESPERANCE,DANIEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: PRATT

DATE/TIME OF TEST: 02/15/2019 22:00:13

Diagnostics Check	OK	21:51	
Air Blank	0.000	21:52	
Control Test	0.079	21:52	
Air Blank	0.000	21:52	
Subject Sample	# 1	21:56	No Sample Provided
Breath Volume	0.000	21:56	
Air Blank	0.000	21:56	
Air Blank	0.000	21:58	
Subject Sample	# 2	21:58	Subject Test Refused
Breath Volume	0.000	21:58	
Air Blank	0.000	21:59	
Control Test	0.080	21:59	
Air Blank	0.000	22:00	
Diagnostics Check	OK	22:00	

Data Download Date/Time: 03/07/2019 09:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LESPERANCE,DANIEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: SEITZ

DATE/TIME OF TEST: 02/16/2019 01:46:27

Diagnostics Check	OK	01:37	
Air Blank	0.000	01:38	
Control Test	0.080	01:38	
Air Blank	0.000	01:39	
Subject Sample	# 1	01:42	No Sample Provided
Breath Volume	0.000	01:42	
Air Blank	0.000	01:42	
Air Blank	0.000	01:44	
Subject Sample	# 2	01:45	Subject Test Refused
Breath Volume	0.000	01:45	
Air Blank	0.000	01:45	
Control Test	0.080	01:45	
Air Blank	0.000	01:46	
Diagnostics Check	OK	01:46	

Data Download Date/Time: 03/07/2019 09:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIGGINBOTHAM,B,K
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: BALLARD,J
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 02/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:11

SUBJECT'S LAST NAME: BURNETTE

DATE/TIME OF TEST: 02/09/2019 20:40:19

Diagnostics Check	OK	20:33
Air Blank	0.000	20:33
Control Test	0.081	20:33
Air Blank	0.000	20:34
Subject Sample # 1	0.164	20:35
Breath Volume	3.718	20:35
Air Blank	0.000	20:36
Air Blank	0.000	20:37
Subject Sample # 2	0.156	20:38
Breath Volume	2.980	20:38
Air Blank	0.000	20:39
Control Test	0.081	20:39
Air Blank	0.000	20:40
Diagnostics Check	OK	20:40

Data Download Date/Time: 02/25/2019 08:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBB,CATHERINE,D
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: ZITTROWER,T
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 02/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: COBB

DATE/TIME OF TEST: 02/10/2019 01:50:23

Diagnostics Check	OK	01:36	
Air Blank	0.000	01:36	
Control Test	0.080	01:36	
Air Blank	0.000	01:37	
Subject Sample	# 1 0.193	01:40	Volume Not Met
Breath Volume	0.148	01:40	
Air Blank	0.000	01:40	
Air Blank	0.000	01:42	
Subject Sample	# 2 0.188	01:45	
Breath Volume	2.203	01:45	
Air Blank	0.000	01:46	
Air Blank	0.000	01:48	
Subject Sample	# 3 0.184	01:48	
Breath Volume	2.324	01:48	
Air Blank	0.000	01:49	
Control Test	0.078	01:49	
Air Blank	0.000	01:50	
Diagnostics Check	OK	01:50	

Data Download Date/Time: 02/25/2019 08:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATHIS,A,J
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: CONE,A
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 02/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: FENNELL

DATE/TIME OF TEST: 02/14/2019 00:26:34

Diagnostics Check	OK	00:23	
Air Blank	0.000	00:24	
Control Test	0.081	00:24	
Air Blank	0.000	00:25	
Subject Sample	# 1	00:25	Subject Test Refused
Breath Volume	0.000	00:25	
Air Blank	0.000	00:25	
Control Test	0.081	00:26	
Air Blank	0.000	00:26	
Diagnostics Check	OK	00:26	

Data Download Date/Time: 02/25/2019 08:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,E,E
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: DELACRUES,C
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 02/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 02/16/2019 03:53:30

Diagnostics Check	OK	03:47
Air Blank	0.000	03:47
Control Test	0.080	03:47
Air Blank	0.000	03:48
Subject Sample # 1	0.154	03:49
Breath Volume	2.140	03:49
Air Blank	0.000	03:49
Air Blank	0.000	03:51
Subject Sample # 2	0.154	03:52
Breath Volume	2.179	03:52
Air Blank	0.000	03:52
Control Test	0.080	03:52
Air Blank	0.000	03:53
Diagnostics Check	OK	03:53

Data Download Date/Time: 02/25/2019 08:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRES,J,L
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 02/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: BINEGAR

DATE/TIME OF TEST: 02/16/2019 04:32:24

Diagnostics Check	OK	04:26
Air Blank	0.000	04:26
Control Test	0.080	04:27
Air Blank	0.000	04:27
Subject Sample # 1	0.128	04:28
Breath Volume	2.363	04:28
Air Blank	0.000	04:28
Air Blank	0.000	04:30
Subject Sample # 2	0.134	04:30
Breath Volume	2.269	04:30
Air Blank	0.000	04:31
Control Test	0.080	04:31
Air Blank	0.000	04:32
Diagnostics Check	OK	04:32

Data Download Date/Time: 02/25/2019 08:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,E,E
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 02/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: CONBOY

DATE/TIME OF TEST: 02/16/2019 23:36:21

Diagnostics Check	OK	23:33	
Air Blank	0.000	23:33	
Control Test	0.080	23:34	
Air Blank	0.000	23:34	
Subject Sample	# 1	23:35	Subject Test Refused
Breath Volume	0.000	23:35	
Air Blank	0.000	23:35	
Control Test	0.080	23:35	
Air Blank	0.000	23:36	
Diagnostics Check	OK	23:36	

Data Download Date/Time: 02/25/2019 08:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,E,E
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 02/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: CONBOY

DATE/TIME OF TEST: 02/17/2019 00:12:54

Diagnostics Check	OK	00:00	
Air Blank	0.000	00:00	
Control Test	0.080	00:01	
Air Blank	0.000	00:01	
Subject Sample	# 1 0.084	00:04	Volume Not Met
Breath Volume	1.078	00:04	
Air Blank	0.000	00:05	
Air Blank	0.000	00:07	
Subject Sample	# 2 0.098	00:08	
Breath Volume	1.773	00:08	
Air Blank	0.000	00:09	
Air Blank	0.000	00:10	
Subject Sample	# 3 0.100	00:11	
Breath Volume	2.824	00:11	
Air Blank	0.000	00:12	
Control Test	0.079	00:12	
Air Blank	0.000	00:12	
Diagnostics Check	OK	00:12	

Data Download Date/Time: 02/25/2019 08:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,E,E
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: TAYLOR,T
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 02/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 02/17/2019 22:41:28

Diagnostics Check	OK	22:34
Air Blank	0.000	22:34
Control Test	0.079	22:35
Air Blank	0.000	22:35
Subject Sample # 1	0.154	22:37
Breath Volume	1.589	22:37
Air Blank	0.000	22:37
Air Blank	0.000	22:39
Subject Sample # 2	0.155	22:40
Breath Volume	1.320	22:40
Air Blank	0.000	22:40
Control Test	0.080	22:40
Air Blank	0.000	22:41
Diagnostics Check	OK	22:41

Data Download Date/Time: 02/25/2019 08:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,D,R
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: REYONLDS,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 02/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: GARTLAND

DATE/TIME OF TEST: 02/18/2019 23:03:47

Diagnostics Check	OK	22:56
Air Blank	0.000	22:56
Control Test	0.079	22:56
Air Blank	0.000	22:57
Subject Sample # 1	0.239	22:58
Breath Volume	2.281	22:58
Air Blank	0.000	22:59
Air Blank	0.000	23:01
Subject Sample # 2	0.234	23:02
Breath Volume	2.960	23:02
Air Blank	0.000	23:02
Control Test	0.078	23:03
Air Blank	0.000	23:03
Diagnostics Check	OK	23:03

Data Download Date/Time: 02/25/2019 08:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCracken,J,D
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 02/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: LUNMAN

DATE/TIME OF TEST: 02/22/2019 00:58:30

Diagnostics Check	OK	00:54	
Air Blank	0.000	00:55	
Control Test	0.080	00:55	
Air Blank	0.000	00:55	
Subject Sample	# 1	00:57	Subject Test Refused
Breath Volume	0.000	00:57	
Air Blank	0.000	00:57	
Control Test	0.080	00:57	
Air Blank	0.000	00:58	
Diagnostics Check	OK	00:58	

Data Download Date/Time: 02/25/2019 08:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,D,R
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: FENDER,A
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 02/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: FOSTER

DATE/TIME OF TEST: 02/23/2019 02:40:00

Diagnostics Check	OK	02:32
Air Blank	0.000	02:33
Control Test	0.080	02:33
Air Blank	0.000	02:34
Subject Sample # 1	0.122	02:35
Breath Volume	1.226	02:35
Air Blank	0.000	02:36
Air Blank	0.000	02:38
Subject Sample # 2	0.116	02:38
Breath Volume	2.007	02:38
Air Blank	0.000	02:39
Control Test	0.080	02:39
Air Blank	0.000	02:39
Diagnostics Check	OK	02:40

Data Download Date/Time: 02/25/2019 08:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBB,CATHERINE,D
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: BALLARD,J
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 02/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: TAPLEY

DATE/TIME OF TEST: 02/24/2019 23:02:22

Diagnostics Check	OK	22:59	
Air Blank	0.000	23:00	
Control Test	0.079	23:00	
Air Blank	0.000	23:00	
Subject Sample	# 1	23:01	Subject Test Refused
Breath Volume	0.000	23:01	
Air Blank	0.000	23:01	
Control Test	0.079	23:01	
Air Blank	0.000	23:02	
Diagnostics Check	OK	23:02	

Data Download Date/Time: 02/25/2019 08:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATHIS,A,J
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: REYNOLDS,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 02/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: HARVEY

DATE/TIME OF TEST: 02/28/2019 04:16:07

Diagnostics Check	OK	04:13	
Air Blank	0.000	04:13	
Control Test	0.080	04:14	
Air Blank	0.000	04:14	
Subject Sample	# 1	04:14	Subject Test Refused
Breath Volume	0.000	04:14	
Air Blank	0.000	04:15	
Control Test	0.080	04:15	
Air Blank	0.000	04:16	
Diagnostics Check	OK	04:16	

Data Download Date/Time: 04/01/2019 16:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBLES,J,V
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: NEW,S
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 01/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: CONLEY

DATE/TIME OF TEST: 02/01/2019 22:56:50

Diagnostics Check	OK	22:53	
Air Blank	0.000	22:54	
Control Test	0.079	22:54	
Air Blank	0.000	22:55	
Subject Sample	# 1	22:55	Subject Test Refused
Breath Volume	0.000	22:55	
Air Blank	0.000	22:55	
Control Test	0.079	22:56	
Air Blank	0.000	22:56	
Diagnostics Check	OK	22:56	

Data Download Date/Time: 02/07/2019 15:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABELN,C,S
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: BUEFORD,S
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 01/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 02/03/2019 04:23:16

Diagnostics Check	OK	04:15	
Air Blank	0.000	04:16	
Control Test	0.079	04:16	
Air Blank	0.000	04:17	
Subject Sample	# 1 0.109	04:18	
Breath Volume	2.472	04:18	
Air Blank	0.000	04:18	
Air Blank	0.000	04:20	
Subject Sample	# 2	04:21	Subject Test Refused
Breath Volume	0.000	04:21	
Air Blank	0.000	04:22	
Control Test	0.079	04:22	
Air Blank	0.000	04:23	
Diagnostics Check	OK	04:23	

Data Download Date/Time: 02/07/2019 15:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABELN,C,S
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: BUEFORD,S
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 01/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 02/03/2019 04:45:32

Diagnostics Check	OK	04:38
Air Blank	0.000	04:39
Control Test	0.079	04:39
Air Blank	0.000	04:40
Subject Sample # 1	0.111	04:41
Breath Volume	2.636	04:41
Air Blank	0.000	04:41
Air Blank	0.000	04:43
Subject Sample # 2	0.111	04:44
Breath Volume	1.597	04:44
Air Blank	0.000	04:44
Control Test	0.079	04:44
Air Blank	0.000	04:45
Diagnostics Check	OK	04:45

Data Download Date/Time: 02/07/2019 15:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROE,ROBERT,D
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: MARIN,ADRIAN
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 01/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: RYAN

DATE/TIME OF TEST: 02/04/2019 03:42:46

Diagnostics Check	OK	03:39	
Air Blank	0.000	03:40	
Control Test	0.080	03:40	
Air Blank	0.000	03:41	
Subject Sample	# 1	03:41	Subject Test Refused
Breath Volume	0.000	03:41	
Air Blank	0.000	03:41	
Control Test	0.079	03:42	
Air Blank	0.000	03:42	
Diagnostics Check	OK	03:42	

Data Download Date/Time: 02/07/2019 15:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROE,ROBERT,D
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SALVADOR,ELDER
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 01/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 02/07/2019 02:29:33

Diagnostics Check	OK	02:26	
Air Blank	0.000	02:27	
Control Test	0.080	02:27	
Air Blank	0.000	02:27	
Subject Sample	# 1	02:28	Subject Test Refused
Breath Volume	0.000	02:28	
Air Blank	0.000	02:28	
Control Test	0.080	02:28	
Air Blank	0.000	02:29	
Diagnostics Check	OK	02:29	

Data Download Date/Time: 02/07/2019 15:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRACKETT,MARK,O
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: BRACKETT,MARK
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 01/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: ZERO TOL
ARREST TIME: 15:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:35

SUBJECT'S LAST NAME: BOGARD

DATE/TIME OF TEST: 02/08/2019 16:06:51

Diagnostics Check	OK	16:00
Air Blank	0.000	16:01
Control Test	0.080	16:01
Air Blank	0.000	16:02
Subject Sample # 1	0.000	16:02
Breath Volume	2.976	16:02
Air Blank	0.000	16:02
Air Blank	0.000	16:04
Subject Sample # 2	0.000	16:05
Breath Volume	1.332	16:05
Air Blank	0.000	16:05
Control Test	0.081	16:06
Air Blank	0.000	16:06
Diagnostics Check	OK	16:06

Data Download Date/Time: 02/20/2019 01:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINISH,BARRY,A
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SALVADOR,ELDER
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 01/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: DELHOMME

DATE/TIME OF TEST: 02/10/2019 04:24:54

Diagnostics Check	OK	04:15
Air Blank	0.000	04:16
Control Test	0.080	04:16
Air Blank	0.000	04:17
Subject Sample # 1	0.272	04:18
Breath Volume	1.527	04:18
Air Blank	0.000	04:19
Air Blank	0.000	04:20
Subject Sample # 2	0.261	04:23
Breath Volume	2.441	04:23
Air Blank	0.000	04:23
Control Test	0.080	04:24
Air Blank	0.000	04:24
Diagnostics Check	OK	04:24

Data Download Date/Time: 02/20/2019 01:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,SUSAN,L
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SALVADOR,E
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 01/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 02/14/2019 03:56:37

Diagnostics Check	OK	03:50
Air Blank	0.000	03:50
Control Test	0.081	03:51
Air Blank	0.000	03:51
Subject Sample # 1	0.038	03:52
Breath Volume	1.941	03:52
Air Blank	0.000	03:52
Air Blank	0.000	03:54
Subject Sample # 2	0.036	03:55
Breath Volume	1.734	03:55
Air Blank	0.000	03:55
Control Test	0.081	03:56
Air Blank	0.000	03:56
Diagnostics Check	OK	03:56

Data Download Date/Time: 02/20/2019 01:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,DUSTIN,J
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: BURRESS,RYAN
ARRESTING OFFICER AGENCY: Lake Placid PD
LAST AGENCY INSPECTION: 01/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:13

SUBJECT'S LAST NAME: CONNOLLY

DATE/TIME OF TEST: 02/17/2019 21:38:26

Diagnostics Check	OK	21:34	
Air Blank	0.000	21:34	
Control Test	0.079	21:35	
Air Blank	0.000	21:35	
Subject Sample	# 1	21:37	Subject Test Refused
Breath Volume	0.000	21:37	
Air Blank	0.000	21:37	
Control Test	0.080	21:37	
Air Blank	0.000	21:38	
Diagnostics Check	OK	21:38	

Data Download Date/Time: 02/20/2019 01:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINISH,BARRY,A
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: HODGE,KYLER
ARRESTING OFFICER AGENCY: Sebring PD
LAST AGENCY INSPECTION: 02/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: TRUJILLO ALVAREZ

DATE/TIME OF TEST: 02/24/2019 01:04:20

Diagnostics Check	OK	00:57
Air Blank	0.000	00:58
Control Test	0.080	00:58
Air Blank	0.000	00:58
Subject Sample # 1	0.209	00:59
Breath Volume	2.203	00:59
Air Blank	0.000	01:00
Air Blank	0.000	01:02
Subject Sample # 2	0.207	01:02
Breath Volume	1.929	01:02
Air Blank	0.000	01:03
Control Test	0.080	01:03
Air Blank	0.000	01:04
Diagnostics Check	OK	01:04

Data Download Date/Time: 03/10/2019 01:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,REBECCA,C
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: RITENOUR,STEVEN
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 02/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:01

SUBJECT'S LAST NAME: KUKAROLA

DATE/TIME OF TEST: 02/25/2019 17:32:49

Diagnostics Check	OK	17:26
Air Blank	0.000	17:26
Control Test	0.080	17:27
Air Blank	0.000	17:27
Subject Sample # 1	0.000	17:28
Breath Volume	1.996	17:28
Air Blank	0.000	17:28
Air Blank	0.000	17:30
Subject Sample # 2	0.000	17:31
Breath Volume	1.574	17:31
Air Blank	0.000	17:31
Control Test	0.081	17:32
Air Blank	0.000	17:32
Diagnostics Check	OK	17:32

Data Download Date/Time: 03/10/2019 01:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,C,L
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: BUEFORD,S
ARRESTING OFFICER AGENCY: Sebring PD
LAST AGENCY INSPECTION: 02/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: BAROUCH

DATE/TIME OF TEST: 02/28/2019 01:13:34

Diagnostics Check	OK	01:10	
Air Blank	0.000	01:11	
Control Test	0.080	01:11	
Air Blank	0.000	01:11	
Subject Sample	# 1	01:12	Subject Test Refused
Breath Volume	0.000	01:12	
Air Blank	0.000	01:12	
Control Test	0.080	01:12	
Air Blank	0.000	01:13	
Diagnostics Check	OK	01:13	

Data Download Date/Time: 03/10/2019 01:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

INSTRUMENT REGISTERED TO: DAVIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JESSUP,BRIAN,C
BREATH TEST OPERATOR AGENCY: Davie PD
ARRESTING OFFICER: PUENTES,YENDI
ARRESTING OFFICER AGENCY: Davie PD
LAST AGENCY INSPECTION: 01/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: TODD

DATE/TIME OF TEST: 02/03/2019 03:31:47

Diagnostics Check	OK	03:25
Air Blank	0.000	03:26
Control Test	0.082	03:26
Air Blank	0.000	03:26
Subject Sample # 1	0.163	03:27
Breath Volume	3.101	03:27
Air Blank	0.000	03:27
Air Blank	0.000	03:29
Subject Sample # 2	0.159	03:30
Breath Volume	2.539	03:30
Air Blank	0.000	03:30
Control Test	0.083	03:31
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time: 02/11/2019 13:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

INSTRUMENT REGISTERED TO: DAVIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STUEHRENBURG, ERIK, C
BREATH TEST OPERATOR AGENCY: Davie PD
ARRESTING OFFICER: STUEHRENBURG, ERIK
ARRESTING OFFICER AGENCY: Davie PD
LAST AGENCY INSPECTION: 02/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: JEANTY

DATE/TIME OF TEST: 02/11/2019 23:05:34

Diagnostics Check	OK	22:58
Air Blank	0.000	22:58
Control Test	0.082	22:59
Air Blank	0.000	22:59
Subject Sample # 1	0.173	23:00
Breath Volume	2.078	23:00
Air Blank	0.000	23:00
Air Blank	0.000	23:02
Subject Sample # 2	0.156	23:03
Breath Volume	2.144	23:03
Air Blank	0.000	23:04
Control Test	0.082	23:04
Air Blank	0.000	23:05
Diagnostics Check	OK	23:05

Data Download Date/Time: 03/06/2019 15:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

INSTRUMENT REGISTERED TO: DAVIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,SHELIA,R
BREATH TEST OPERATOR AGENCY: Davie PD
ARRESTING OFFICER: FERNANDEZ,KENNY
ARRESTING OFFICER AGENCY: Davie PD
LAST AGENCY INSPECTION: 02/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 17:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:32

SUBJECT'S LAST NAME: AGUILA

DATE/TIME OF TEST: 02/19/2019 19:03:57

Diagnostics Check	OK	18:57
Air Blank	0.000	18:57
Control Test	0.081	18:58
Air Blank	0.000	18:58
Subject Sample # 1	0.218	18:59
Breath Volume	1.289	18:59
Air Blank	0.000	19:00
Air Blank	0.000	19:01
Subject Sample # 2	0.216	19:02
Breath Volume	2.636	19:02
Air Blank	0.000	19:03
Control Test	0.080	19:03
Air Blank	0.000	19:03
Diagnostics Check	OK	19:03

Data Download Date/Time: 03/06/2019 15:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

INSTRUMENT REGISTERED TO: DAVIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAHAM,HUGH,E
BREATH TEST OPERATOR AGENCY: Davie PD
ARRESTING OFFICER: SABARESE,ANTHONY
ARRESTING OFFICER AGENCY: Davie PD
LAST AGENCY INSPECTION: 02/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 06:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:35

SUBJECT'S LAST NAME: ALI

DATE/TIME OF TEST: 02/24/2019 07:07:37

Diagnostics Check	OK	07:01
Air Blank	0.000	07:02
Control Test	0.082	07:02
Air Blank	0.000	07:02
Subject Sample # 1	0.079	07:03
Breath Volume	1.253	07:03
Air Blank	0.000	07:03
Air Blank	0.000	07:05
Subject Sample # 2	0.091	07:06
Breath Volume	2.953	07:06
Air Blank	0.000	07:06
Control Test	0.081	07:07
Air Blank	0.000	07:07
Diagnostics Check	OK	07:07

Data Download Date/Time: 03/06/2019 15:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALMEIDA,CAMILO,J
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: ALMEIDA,CAMILO
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 01/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: HODGES

DATE/TIME OF TEST: 02/08/2019 00:37:51

Diagnostics Check	OK	00:31
Air Blank	0.000	00:31
Control Test	0.081	00:32
Air Blank	0.000	00:32
Subject Sample # 1	0.151	00:33
Breath Volume	1.281	00:33
Air Blank	0.000	00:33
Air Blank	0.000	00:35
Subject Sample # 2	0.163	00:36
Breath Volume	2.332	00:36
Air Blank	0.000	00:36
Control Test	0.080	00:37
Air Blank	0.000	00:37
Diagnostics Check	OK	00:37

Data Download Date/Time: 02/18/2019 11:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SUTTLE,CORY,A

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: SUTTLE,CORY

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 01/14/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17817080A3

DRY GAS EXPIRY DATE: 08/05/2019

VIOLATION CODE: DUI

ARREST TIME: 02:15

GENDER: Female

OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: STEPHENS

DATE/TIME OF TEST: 02/14/2019 04:12:50

Diagnostics Check	OK	04:08	
Air Blank	0.000	04:09	
Control Test	0.080	04:09	
Air Blank	0.000	04:09	
Subject Sample	# 1	04:11	Subject Test Refused
Breath Volume	0.000	04:11	
Air Blank	0.000	04:11	
Control Test	0.081	04:12	
Air Blank	0.000	04:12	
Diagnostics Check	OK	04:12	

Data Download Date/Time: 02/18/2019 11:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTLE,CORY,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SUTTLE,CORY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 01/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 02/15/2019 03:26:52

Diagnostics Check	OK	03:21
Air Blank	0.000	03:21
Control Test	0.079	03:21
Air Blank	0.000	03:22
Subject Sample # 1	0.000	03:22
Breath Volume	2.265	03:22
Air Blank	0.000	03:23
Air Blank	0.000	03:25
Subject Sample # 2	0.000	03:25
Breath Volume	2.550	03:25
Air Blank	0.000	03:25
Control Test	0.079	03:26
Air Blank	0.000	03:26
Diagnostics Check	OK	03:26

Data Download Date/Time: 02/18/2019 11:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALMEIDA,CAMILO,J
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SUTTLE,CORY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 01/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: FLORES

DATE/TIME OF TEST: 02/18/2019 03:49:44

Diagnostics Check	OK	03:40
Air Blank	0.000	03:41
Control Test	0.079	03:41
Air Blank	0.000	03:42
Subject Sample # 1	0.203	03:42
Breath Volume	1.386	03:42
Air Blank	0.000	03:43
Air Blank	0.000	03:45
Subject Sample # 2	0.215	03:48
Breath Volume	1.683	03:48
Air Blank	0.000	03:48
Control Test	0.079	03:49
Air Blank	0.000	03:49
Diagnostics Check	OK	03:49

Data Download Date/Time: 02/18/2019 11:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULKEARNS,MICHAEL,S
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: MERRITT

DATE/TIME OF TEST: 02/02/2019 23:25:55

Diagnostics Check	OK	23:19
Air Blank	0.000	23:19
Control Test	0.078	23:20
Air Blank	0.000	23:20
Subject Sample # 1	0.181	23:21
Breath Volume	2.457	23:21
Air Blank	0.000	23:21
Air Blank	0.000	23:23
Subject Sample # 2	0.186	23:24
Breath Volume	2.375	23:24
Air Blank	0.000	23:24
Control Test	0.079	23:25
Air Blank	0.000	23:25
Diagnostics Check	OK	23:25

Data Download Date/Time: 02/13/2019 10:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULKEARNS,MICHAEL,S
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: DELACRUZ,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: VILLA

DATE/TIME OF TEST: 02/02/2019 23:45:01

Diagnostics Check	OK	23:42	
Air Blank	0.000	23:42	
Control Test	0.078	23:42	
Air Blank	0.000	23:43	
Subject Sample	# 1	23:43	Subject Test Refused
Breath Volume	0.000	23:43	
Air Blank	0.000	23:44	
Control Test	0.077	23:44	
Air Blank	0.000	23:44	
Diagnostics Check	OK	23:45	

Data Download Date/Time: 02/13/2019 10:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULKEARNS,MICHAEL,S
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:26

SUBJECT'S LAST NAME: BROWN JR

DATE/TIME OF TEST: 02/03/2019 21:51:33

Diagnostics Check	OK	21:48	
Air Blank	0.000	21:49	
Control Test	0.077	21:49	
Air Blank	0.000	21:50	
Subject Sample	# 1	21:50	Subject Test Refused
Breath Volume	0.000	21:50	
Air Blank	0.000	21:50	
Control Test	0.077	21:50	
Air Blank	0.000	21:51	
Diagnostics Check	OK	21:51	

Data Download Date/Time: 02/13/2019 10:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULKEARNS,MICHAEL,S
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 01/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: LINDON

DATE/TIME OF TEST: 02/04/2019 01:54:22

Diagnostics Check	OK	01:47
Air Blank	0.000	01:47
Control Test	0.077	01:48
Air Blank	0.000	01:48
Subject Sample # 1	0.208	01:49
Breath Volume	1.781	01:49
Air Blank	0.000	01:49
Air Blank	0.000	01:51
Subject Sample # 2	0.193	01:52
Breath Volume	1.871	01:52
Air Blank	0.000	01:53
Control Test	0.078	01:53
Air Blank	0.000	01:54
Diagnostics Check	OK	01:54

Data Download Date/Time: 02/13/2019 10:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

INSTRUMENT REGISTERED TO: WINTER HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORPHEE,MICHAEL,
BREATH TEST OPERATOR AGENCY: Winter Haven PD
ARRESTING OFFICER: CRUZ,EDGARDO
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 01/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: MC CASLIN

DATE/TIME OF TEST: 02/05/2019 23:55:35

Diagnostics Check	OK	23:48
Air Blank	0.000	23:49
Control Test	0.078	23:49
Air Blank	0.000	23:50
Subject Sample # 1	0.153	23:50
Breath Volume	2.054	23:50
Air Blank	0.000	23:51
Air Blank	0.000	23:53
Subject Sample # 2	0.147	23:53
Breath Volume	1.875	23:53
Air Blank	0.000	23:54
Control Test	0.079	23:54
Air Blank	0.000	23:55
Diagnostics Check	OK	23:55

Data Download Date/Time: 02/27/2019 15:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

INSTRUMENT REGISTERED TO: WINTER HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COAKE,SPENCER,M
BREATH TEST OPERATOR AGENCY: Winter Haven PD
ARRESTING OFFICER: CRUZ,EDGARDO
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 01/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: ILLESCA-ESPINOSA

DATE/TIME OF TEST: 02/09/2019 05:01:55

Diagnostics Check	OK	04:54
Air Blank	0.000	04:55
Control Test	0.079	04:55
Air Blank	0.000	04:56
Subject Sample # 1	0.166	04:56
Breath Volume	1.339	04:56
Air Blank	0.000	04:57
Air Blank	0.000	04:59
Subject Sample # 2	0.173	05:00
Breath Volume	3.019	05:00
Air Blank	0.000	05:00
Control Test	0.083	05:01
Air Blank	0.000	05:01
Diagnostics Check	OK	05:01

Data Download Date/Time: 02/27/2019 15:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

INSTRUMENT REGISTERED TO: WINTER HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOMACK,ERIC,V
BREATH TEST OPERATOR AGENCY: Winter Haven PD
ARRESTING OFFICER: BOONE,BRANDON
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 01/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: COURT ORD
ARREST TIME: 09:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:15

SUBJECT'S LAST NAME: ABBOTT

DATE/TIME OF TEST: 02/12/2019 09:42:39

Diagnostics Check	OK	09:36
Air Blank	0.000	09:36
Control Test	0.079	09:37
Air Blank	0.000	09:37
Subject Sample # 1	0.000	09:38
Breath Volume	2.394	09:38
Air Blank	0.000	09:38
Air Blank	0.000	09:40
Subject Sample # 2	0.000	09:41
Breath Volume	2.605	09:41
Air Blank	0.000	09:41
Control Test	0.081	09:41
Air Blank	0.000	09:42
Diagnostics Check	OK	09:42

Data Download Date/Time: 02/27/2019 15:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

INSTRUMENT REGISTERED TO: WINTER HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NICHOLS,KENNETH,D
BREATH TEST OPERATOR AGENCY: Winter Haven PD
ARRESTING OFFICER: NICHOLS,KENNETH
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 01/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: PROB/PAROL
ARREST TIME: 16:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 02/13/2019 16:58:52

Diagnostics Check	OK	16:52
Air Blank	0.000	16:53
Control Test	0.079	16:53
Air Blank	0.000	16:54
Subject Sample # 1	0.000	16:54
Breath Volume	2.253	16:54
Air Blank	0.000	16:54
Air Blank	0.000	16:56
Subject Sample # 2	0.000	16:57
Breath Volume	2.347	16:57
Air Blank	0.000	16:57
Control Test	0.080	16:58
Air Blank	0.000	16:58
Diagnostics Check	OK	16:58

Data Download Date/Time: 02/27/2019 15:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

INSTRUMENT REGISTERED TO: WINTER HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEBSON,BRET,E
BREATH TEST OPERATOR AGENCY: Winter Haven PD
ARRESTING OFFICER: HEBSON,BRET
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 01/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: COURT ORD
ARREST TIME: 09:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:34

SUBJECT'S LAST NAME: LARSON

DATE/TIME OF TEST: 02/15/2019 10:13:06

Diagnostics Check	OK	10:07
Air Blank	0.000	10:07
Control Test	0.079	10:07
Air Blank	0.000	10:08
Subject Sample # 1	0.000	10:08
Breath Volume	2.421	10:08
Air Blank	0.000	10:09
Air Blank	0.000	10:11
Subject Sample # 2	0.000	10:11
Breath Volume	2.230	10:11
Air Blank	0.000	10:12
Control Test	0.080	10:12
Air Blank	0.000	10:13
Diagnostics Check	OK	10:13

Data Download Date/Time: 02/27/2019 15:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

INSTRUMENT REGISTERED TO: WINTER HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIERZBA,NATHAN,J
BREATH TEST OPERATOR AGENCY: Winter Haven PD
ARRESTING OFFICER: WIERZBA,NATHAN
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 01/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: WASHINGTON

DATE/TIME OF TEST: 02/20/2019 21:24:12

Diagnostics Check	OK	21:21	
Air Blank	0.000	21:21	
Control Test	0.078	21:22	
Air Blank	0.000	21:22	
Subject Sample	# 1	21:22	Subject Test Refused
Breath Volume	0.000	21:22	
Air Blank	0.000	21:23	
Control Test	0.078	21:23	
Air Blank	0.000	21:24	
Diagnostics Check	OK	21:24	

Data Download Date/Time: 02/27/2019 15:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

INSTRUMENT REGISTERED TO: WINTER HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOMACK,ERIC,V
BREATH TEST OPERATOR AGENCY: Winter Haven PD
ARRESTING OFFICER: WOMACK,ERIC
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 01/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: COURT ORD
ARREST TIME: 08:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:04

SUBJECT'S LAST NAME: LARSON

DATE/TIME OF TEST: 02/21/2019 09:30:32

Diagnostics Check	OK	09:24
Air Blank	0.000	09:24
Control Test	0.079	09:25
Air Blank	0.000	09:25
Subject Sample # 1	0.000	09:26
Breath Volume	2.339	09:26
Air Blank	0.000	09:26
Air Blank	0.000	09:28
Subject Sample # 2	0.000	09:29
Breath Volume	2.531	09:29
Air Blank	0.000	09:29
Control Test	0.079	09:29
Air Blank	0.000	09:30
Diagnostics Check	OK	09:30

Data Download Date/Time: 02/27/2019 15:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

INSTRUMENT REGISTERED TO: WINTER HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOMACK,ERIC,V
BREATH TEST OPERATOR AGENCY: Winter Haven PD
ARRESTING OFFICER: WOMACK,ERIC
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 02/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: COURT ORD
ARREST TIME: 08:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:54

SUBJECT'S LAST NAME: HERMAN

DATE/TIME OF TEST: 02/28/2019 09:20:46

Diagnostics Check	OK	09:14
Air Blank	0.000	09:14
Control Test	0.079	09:15
Air Blank	0.000	09:15
Subject Sample # 1	0.000	09:16
Breath Volume	2.285	09:16
Air Blank	0.000	09:16
Air Blank	0.000	09:18
Subject Sample # 2	0.000	09:19
Breath Volume	3.156	09:19
Air Blank	0.000	09:19
Control Test	0.080	09:20
Air Blank	0.000	09:20
Diagnostics Check	OK	09:20

Data Download Date/Time: 03/27/2019 10:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

INSTRUMENT REGISTERED TO: HOLLYWOOD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DERIET,RAFAEL,
BREATH TEST OPERATOR AGENCY: Hollywood PD
ARRESTING OFFICER: MARINO,GREG
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 01/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 07:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:45

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 02/21/2019 08:12:40

Diagnostics Check	OK	08:06
Air Blank	0.000	08:06
Control Test	0.079	08:07
Air Blank	0.000	08:07
Subject Sample # 1	0.235	08:08
Breath Volume	2.246	08:08
Air Blank	0.000	08:08
Air Blank	0.000	08:10
Subject Sample # 2	0.241	08:11
Breath Volume	3.222	08:11
Air Blank	0.000	08:11
Control Test	0.079	08:12
Air Blank	0.000	08:12
Diagnostics Check	OK	08:12

Data Download Date/Time: 02/21/2019 12:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAZIME,OFC. A.,H
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: HAZIME,OFC. A.
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 01/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 02/03/2019 04:54:32

Diagnostics Check	OK	04:48
Air Blank	0.000	04:48
Control Test	0.079	04:48
Air Blank	0.000	04:49
Subject Sample # 1	0.179	04:49
Breath Volume	2.929	04:49
Air Blank	0.000	04:50
Air Blank	0.000	04:52
Subject Sample # 2	0.181	04:53
Breath Volume	2.703	04:53
Air Blank	0.000	04:53
Control Test	0.079	04:53
Air Blank	0.000	04:54
Diagnostics Check	OK	04:54

Data Download Date/Time: 02/26/2019 22:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAZIME,OFC. A.,
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: HAZIME,OFC. A.
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 01/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 02/12/2019 03:12:22

Diagnostics Check	OK	03:05
Air Blank	0.000	03:06
Control Test	0.079	03:06
Air Blank	0.000	03:06
Subject Sample # 1	0.154	03:07
Breath Volume	3.515	03:07
Air Blank	0.000	03:08
Air Blank	0.000	03:10
Subject Sample # 2	0.166	03:10
Breath Volume	3.203	03:10
Air Blank	0.000	03:11
Control Test	0.078	03:11
Air Blank	0.000	03:12
Diagnostics Check	OK	03:12

Data Download Date/Time: 02/26/2019 22:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAZIME,OFC. A.,
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: SABO,OFC. T.
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 01/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: DOKKEN

DATE/TIME OF TEST: 02/12/2019 23:40:09

Diagnostics Check	OK	23:32
Air Blank	0.000	23:33
Control Test	0.079	23:33
Air Blank	0.000	23:34
Subject Sample # 1	0.211	23:35
Breath Volume	1.324	23:35
Air Blank	0.000	23:36
Air Blank	0.000	23:38
Subject Sample # 2	0.222	23:38
Breath Volume	3.160	23:38
Air Blank	0.000	23:39
Control Test	0.078	23:39
Air Blank	0.000	23:40
Diagnostics Check	OK	23:40

Data Download Date/Time: 02/26/2019 22:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAZIME,OFC. A.,
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: HAZIME,OFC. A.
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 01/15/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: REAGAN

DATE/TIME OF TEST: 02/17/2019 00:38:39

Diagnostics Check	OK	00:32
Air Blank	0.000	00:32
Control Test	0.079	00:32
Air Blank	0.000	00:33
Subject Sample # 1	0.132	00:33
Breath Volume	2.570	00:33
Air Blank	0.000	00:34
Air Blank	0.000	00:36
Subject Sample # 2	0.127	00:37
Breath Volume	2.433	00:37
Air Blank	0.000	00:37
Control Test	0.079	00:38
Air Blank	0.000	00:38
Diagnostics Check	OK	00:38

Data Download Date/Time: 02/26/2019 22:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAZIME,OFC. A.,
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: HAZIME,OFC. A.
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 01/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: KUIITE

DATE/TIME OF TEST: 02/21/2019 00:21:30

Diagnostics Check	OK	00:14
Air Blank	0.000	00:15
Control Test	0.078	00:15
Air Blank	0.000	00:16
Subject Sample # 1	0.108	00:16
Breath Volume	3.261	00:16
Air Blank	0.000	00:17
Air Blank	0.000	00:19
Subject Sample # 2	0.096	00:19
Breath Volume	1.742	00:19
Air Blank	0.000	00:20
Control Test	0.079	00:20
Air Blank	0.000	00:21
Diagnostics Check	OK	00:21

Data Download Date/Time: 02/26/2019 22:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAZIME,OFC. A.,
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: HAZIME,OFC. A.
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 02/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: TOFT

DATE/TIME OF TEST: 02/27/2019 01:59:56

Diagnostics Check	OK	01:53
Air Blank	0.000	01:54
Control Test	0.080	01:54
Air Blank	0.000	01:54
Subject Sample # 1	0.010	01:55
Breath Volume	3.710	01:55
Air Blank	0.000	01:55
Air Blank	0.000	01:57
Subject Sample # 2	0.010	01:58
Breath Volume	3.992	01:58
Air Blank	0.000	01:59
Control Test	0.080	01:59
Air Blank	0.000	01:59
Diagnostics Check	OK	01:59

Data Download Date/Time: 03/26/2019 22:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RONDOS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: RONDOS,ROBERT
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 01/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:12

SUBJECT'S LAST NAME: GLOWATY

DATE/TIME OF TEST: 02/01/2019 21:42:02

Diagnostics Check	OK	21:38	
Air Blank	0.000	21:39	
Control Test	0.080	21:39	
Air Blank	0.000	21:40	
Subject Sample	# 1	21:40	Subject Test Refused
Breath Volume	0.000	21:40	
Air Blank	0.000	21:41	
Control Test	0.080	21:41	
Air Blank	0.000	21:41	
Diagnostics Check	OK	21:42	

Data Download Date/Time: 02/26/2019 22:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RONDOS,ROBERT,E
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: RONDOS,ROBERT
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 01/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:37

SUBJECT'S LAST NAME: SANZ PEREZ

DATE/TIME OF TEST: 02/06/2019 22:01:46

Diagnostics Check	OK	21:57	
Air Blank	0.000	21:58	
Control Test	0.081	21:58	
Air Blank	0.000	21:59	
Subject Sample	# 1	22:00	Subject Test Refused
Breath Volume	0.000	22:00	
Air Blank	0.000	22:00	
Control Test	0.081	22:01	
Air Blank	0.000	22:01	
Diagnostics Check	OK	22:01	

Data Download Date/Time: 02/26/2019 22:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,JACOB,R
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: MILLER,JACOB
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 01/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:09

SUBJECT'S LAST NAME: HICKMAN

DATE/TIME OF TEST: 02/06/2019 22:32:38

Diagnostics Check	OK	22:29	
Air Blank	0.000	22:30	
Control Test	0.081	22:30	
Air Blank	0.000	22:31	
Subject Sample	# 1	22:31	Subject Test Refused
Breath Volume	0.000	22:31	
Air Blank	0.000	22:31	
Control Test	0.081	22:32	
Air Blank	0.000	22:32	
Diagnostics Check	OK	22:32	

Data Download Date/Time: 02/26/2019 22:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,JACOB,R
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: MILLER,JACOB
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 01/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: MELITO JR

DATE/TIME OF TEST: 02/17/2019 00:58:32

Diagnostics Check	OK	00:46	
Air Blank	0.000	00:46	
Control Test	0.081	00:47	
Air Blank	0.000	00:47	
Subject Sample	# 1	00:50	No Sample Provided
Breath Volume	0.000	00:50	
Air Blank	0.000	00:51	
Air Blank	0.000	00:53	
Subject Sample	# 2	00:53	
Breath Volume	3.625	00:53	
Air Blank	0.000	00:54	
Air Blank	0.000	00:56	
Subject Sample	# 3	00:57	
Breath Volume	1.785	00:57	
Air Blank	0.000	00:57	
Control Test	0.081	00:57	
Air Blank	0.000	00:58	
Diagnostics Check	OK	00:58	

Data Download Date/Time: 02/26/2019 22:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEDE,BRENNEN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WEDE,BRENNEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: DIAZ-MENDEZ

DATE/TIME OF TEST: 02/03/2019 01:50:37

Diagnostics Check	OK	01:37	
Air Blank	0.000	01:38	
Control Test	0.080	01:38	
Air Blank	0.000	01:39	
Subject Sample	# 1 0.185	01:42	Volume Not Met
Breath Volume	0.757	01:42	
Air Blank	0.000	01:42	
Air Blank	0.000	01:44	
Subject Sample	# 2 0.184	01:45	
Breath Volume	1.613	01:45	
Air Blank	0.000	01:46	
Air Blank	0.000	01:48	
Subject Sample	# 3 0.192	01:49	
Breath Volume	1.925	01:49	
Air Blank	0.000	01:49	
Control Test	0.079	01:50	
Air Blank	0.000	01:50	
Diagnostics Check	OK	01:50	

Data Download Date/Time: 03/20/2019 23:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 02/03/2019 01:58:54

Diagnostics Check	OK	01:52
Air Blank	0.000	01:52
Control Test	0.080	01:53
Air Blank	0.000	01:53
Subject Sample # 1	0.282	01:54
Breath Volume	3.093	01:54
Air Blank	0.000	01:54
Air Blank	0.000	01:56
Subject Sample # 2	0.274	01:57
Breath Volume	2.503	01:57
Air Blank	0.000	01:57
Control Test	0.080	01:58
Air Blank	0.000	01:58
Diagnostics Check	OK	01:58

Data Download Date/Time: 03/20/2019 23:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: STOLA

DATE/TIME OF TEST: 02/03/2019 03:46:22

Diagnostics Check	OK	03:40
Air Blank	0.000	03:40
Control Test	0.081	03:41
Air Blank	0.000	03:41
Subject Sample # 1	0.190	03:42
Breath Volume	2.726	03:42
Air Blank	0.000	03:42
Air Blank	0.000	03:44
Subject Sample # 2	0.183	03:44
Breath Volume	3.160	03:44
Air Blank	0.000	03:45
Control Test	0.081	03:45
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

Data Download Date/Time: 03/20/2019 23:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 02/03/2019 04:30:45

Diagnostics Check	OK	04:22
Air Blank	0.000	04:22
Control Test	0.081	04:23
Air Blank	0.000	04:23
Subject Sample # 1	0.098	04:25
Breath Volume	1.191	04:25
Air Blank	0.000	04:25
Air Blank	0.000	04:27
Subject Sample # 2	0.095	04:29
Breath Volume	1.308	04:29
Air Blank	0.000	04:29
Control Test	0.081	04:30
Air Blank	0.000	04:30
Diagnostics Check	OK	04:30

Data Download Date/Time: 03/20/2019 23:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RANISZEWSKI,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: SCHRADER

DATE/TIME OF TEST: 02/04/2019 01:52:34

Diagnostics Check	OK	01:43	
Air Blank	0.000	01:43	
Control Test	0.081	01:44	
Air Blank	0.000	01:44	
Subject Sample	# 1	01:47	No Sample Provided
Breath Volume	0.000	01:47	
Air Blank	0.000	01:48	
Air Blank	0.000	01:50	
Subject Sample	# 2	01:51	Subject Test Refused
Breath Volume	0.000	01:51	
Air Blank	0.000	01:51	
Control Test	0.081	01:51	
Air Blank	0.000	01:52	
Diagnostics Check	OK	01:52	

Data Download Date/Time: 03/20/2019 23:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RANISZEWSKI,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: SCHRADER

DATE/TIME OF TEST: 02/04/2019 02:06:48

Diagnostics Check	OK	01:57
Air Blank	0.000	01:58
Control Test	0.081	01:58
Air Blank	0.000	01:58
Subject Sample # 1	0.171	02:01
Breath Volume	1.144	02:01
Air Blank	0.000	02:01
Air Blank	0.000	02:03
Subject Sample # 2	0.168	02:05
Breath Volume	1.179	02:05
Air Blank	0.000	02:05
Control Test	0.080	02:06
Air Blank	0.000	02:06
Diagnostics Check	OK	02:06

Data Download Date/Time: 03/20/2019 23:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MITCHELL,R
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: PAUL

DATE/TIME OF TEST: 02/04/2019 03:48:58

Diagnostics Check	OK	03:45	
Air Blank	0.000	03:46	
Control Test	0.080	03:46	
Air Blank	0.000	03:47	
Subject Sample	# 1	03:47	Subject Test Refused
Breath Volume	0.000	03:47	
Air Blank	0.000	03:48	
Control Test	0.081	03:48	
Air Blank	0.000	03:48	
Diagnostics Check	OK	03:48	

Data Download Date/Time: 03/20/2019 23:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REEVES, MARK, O
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HANSEN, DUSTIN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:37

SUBJECT'S LAST NAME: FIELDS

DATE/TIME OF TEST: 02/05/2019 16:05:13

Diagnostics Check	OK	15:58	
Air Blank	0.000	15:58	
Control Test	0.082	15:58	
Air Blank	0.000	15:59	
Subject Sample	# 1 0.237	16:00	
Breath Volume	3.421	16:00	
Air Blank	0.000	16:00	
Air Blank	0.000	16:02	
Subject Sample	# 2 0.246	16:03	
Breath Volume	3.089	16:03	
Air Blank	0.000	16:04	
Control Test	0.080	16:04	
Air Blank		16:04	RFI Detect
Air Blank	0.000	16:05	

Data Download Date/Time: 03/20/2019 23:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REEVES, MARK, O
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HANSEN, DUSTIN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:40

SUBJECT'S LAST NAME: FIELDS

DATE/TIME OF TEST: 02/05/2019 16:13:58

Diagnostics Check	OK	16:07
Air Blank	0.000	16:07
Control Test	0.079	16:08
Air Blank	0.000	16:08
Subject Sample # 1	0.242	16:09
Breath Volume	3.382	16:09
Air Blank	0.000	16:09
Air Blank	0.000	16:11
Subject Sample # 2	0.234	16:12
Breath Volume	2.660	16:12
Air Blank	0.000	16:13
Control Test	0.080	16:13
Air Blank	0.000	16:13
Diagnostics Check	OK	16:13

Data Download Date/Time: 03/20/2019 23:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORNING,PETER,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HARLAN,PATRICK
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 11:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:48

SUBJECT'S LAST NAME: PATTBERG

DATE/TIME OF TEST: 02/06/2019 13:20:47

Diagnostics Check	OK	13:08	
Air Blank	0.000	13:08	
Control Test	0.082	13:09	
Air Blank	0.000	13:09	
Subject Sample	# 1 0.030	13:12	Volume Not Met
Breath Volume	0.300	13:12	
Air Blank	0.000	13:13	
Air Blank	0.000	13:15	
Subject Sample	# 2 0.028	13:15	
Breath Volume	1.222	13:15	
Air Blank	0.000	13:16	
Air Blank	0.000	13:18	
Subject Sample	# 3 0.027	13:19	
Breath Volume	1.289	13:19	
Air Blank	0.000	13:19	
Control Test	0.080	13:20	
Air Blank	0.000	13:20	
Diagnostics Check	OK	13:20	

Data Download Date/Time: 03/20/2019 23:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LEFFINGWELL,B
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: PADGETT

DATE/TIME OF TEST: 02/06/2019 22:29:50

Diagnostics Check	OK	22:25	
Air Blank	0.000	22:25	
Control Test	0.081	22:26	
Air Blank	0.000	22:26	
Subject Sample	# 1	22:28	Subject Test Refused
Breath Volume	0.000	22:28	
Air Blank	0.000	22:28	
Control Test	0.081	22:29	
Air Blank	0.000	22:29	
Diagnostics Check	OK	22:29	

Data Download Date/Time: 03/20/2019 23:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RANISZEWSKI,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: YARBROUGH

DATE/TIME OF TEST: 02/07/2019 01:23:16

Diagnostics Check	OK	01:13
Air Blank	0.000	01:13
Control Test	0.081	01:13
Air Blank	0.000	01:14
Subject Sample # 1	0.315	01:15
Breath Volume	1.679	01:15
Air Blank	0.000	01:15
Air Blank	0.000	01:17
Subject Sample # 2	0.290	01:18
Breath Volume	1.304	01:18
Air Blank	0.000	01:19
Air Blank	0.000	01:20
Subject Sample # 3	0.324	01:21
Breath Volume	1.464	01:21
Air Blank	0.000	01:22
Control Test	0.080	01:22
Air Blank	0.000	01:23
Diagnostics Check	OK	01:23

Data Download Date/Time: 03/20/2019 23:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BEHRENSHOUSER,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: KRSTANOSKI-TODROW

DATE/TIME OF TEST: 02/07/2019 02:37:58

Diagnostics Check	OK	02:28
Air Blank	0.000	02:28
Control Test	0.081	02:29
Air Blank	0.000	02:29
Subject Sample # 1	0.000	02:32
Breath Volume	1.355	02:32
Air Blank	0.000	02:33
Air Blank	0.000	02:35
Subject Sample # 2	0.000	02:36
Breath Volume	1.285	02:36
Air Blank	0.000	02:37
Control Test	0.080	02:37
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

Data Download Date/Time: 03/20/2019 23:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: DEJESUS JR

DATE/TIME OF TEST: 02/07/2019 03:37:46

Diagnostics Check	OK	03:31
Air Blank	0.000	03:31
Control Test	0.081	03:32
Air Blank	0.000	03:32
Subject Sample # 1	0.084	03:33
Breath Volume	1.109	03:33
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.085	03:36
Breath Volume	2.339	03:36
Air Blank	0.000	03:36
Control Test	0.081	03:37
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 03/20/2019 23:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICOTTILLI,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI,JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: SIM

DATE/TIME OF TEST: 02/07/2019 05:00:08

Diagnostics Check	OK	04:52
Air Blank	0.000	04:53
Control Test	0.081	04:53
Air Blank	0.000	04:54
Subject Sample # 1	0.156	04:55
Breath Volume	2.000	04:55
Air Blank	0.000	04:55
Air Blank	0.000	04:57
Subject Sample # 2	0.148	04:58
Breath Volume	1.386	04:58
Air Blank	0.000	04:59
Control Test	0.081	04:59
Air Blank	0.000	05:00
Diagnostics Check	OK	05:00

Data Download Date/Time: 03/20/2019 23:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEDE,BRENNEN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WEDE,BRENNEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: MCCOY

DATE/TIME OF TEST: 02/07/2019 23:55:12

Diagnostics Check	OK	23:48
Air Blank	0.000	23:49
Control Test	0.081	23:49
Air Blank	0.000	23:50
Subject Sample # 1	0.218	23:50
Breath Volume	1.457	23:50
Air Blank	0.000	23:51
Air Blank	0.000	23:53
Subject Sample # 2	0.198	23:53
Breath Volume	1.847	23:53
Air Blank	0.000	23:54
Control Test	0.081	23:54
Air Blank	0.000	23:55
Diagnostics Check	OK	23:55

Data Download Date/Time: 03/20/2019 23:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: MUNGUIA AVILA

DATE/TIME OF TEST: 02/08/2019 02:38:19

Diagnostics Check	OK	02:35	
Air Blank	0.000	02:35	
Control Test	0.082	02:36	
Air Blank	0.000	02:36	
Subject Sample	# 1	02:36	Subject Test Refused
Breath Volume	0.000	02:36	
Air Blank	0.000	02:37	
Control Test	0.082	02:37	
Air Blank	0.000	02:38	
Diagnostics Check	OK	02:38	

Data Download Date/Time: 03/20/2019 23:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: ACIERNO

DATE/TIME OF TEST: 02/08/2019 04:37:35

Diagnostics Check	OK	04:31
Air Blank	0.000	04:31
Control Test	0.082	04:32
Air Blank	0.000	04:32
Subject Sample # 1	0.191	04:33
Breath Volume	3.144	04:33
Air Blank	0.000	04:33
Air Blank	0.000	04:35
Subject Sample # 2	0.177	04:36
Breath Volume	3.832	04:36
Air Blank	0.000	04:36
Control Test	0.080	04:37
Air Blank	0.000	04:37
Diagnostics Check	OK	04:37

Data Download Date/Time: 03/20/2019 23:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: DUFFY

DATE/TIME OF TEST: 02/09/2019 00:23:12

Diagnostics Check	OK	00:20	
Air Blank	0.000	00:20	
Control Test	0.081	00:21	
Air Blank	0.000	00:21	
Subject Sample	# 1	00:21	Subject Test Refused
Breath Volume	0.000	00:21	
Air Blank	0.000	00:22	
Control Test	0.081	00:22	
Air Blank	0.000	00:23	
Diagnostics Check	OK	00:23	

Data Download Date/Time: 03/20/2019 23:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PUPKE,L
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: TEWELL

DATE/TIME OF TEST: 02/09/2019 02:16:57

Diagnostics Check	OK	02:10
Air Blank	0.000	02:10
Control Test	0.082	02:11
Air Blank	0.000	02:11
Subject Sample # 1	0.186	02:12
Breath Volume	2.253	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:14
Subject Sample # 2	0.193	02:15
Breath Volume	1.746	02:15
Air Blank	0.000	02:16
Control Test	0.080	02:16
Air Blank	0.000	02:16
Diagnostics Check	OK	02:16

Data Download Date/Time: 03/20/2019 23:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,L
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 02/09/2019 02:26:28

Diagnostics Check	OK	02:23	
Air Blank	0.000	02:24	
Control Test	0.081	02:24	
Air Blank	0.000	02:24	
Subject Sample	# 1	02:25	Subject Test Refused
Breath Volume	0.000	02:25	
Air Blank	0.000	02:25	
Control Test	0.081	02:25	
Air Blank	0.000	02:26	
Diagnostics Check	OK	02:26	

Data Download Date/Time: 03/20/2019 23:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: BOONE

DATE/TIME OF TEST: 02/09/2019 03:06:16

Diagnostics Check	OK	02:59
Air Blank	0.000	02:59
Control Test	0.081	02:59
Air Blank	0.000	03:00
Subject Sample # 1	0.142	03:01
Breath Volume	1.152	03:01
Air Blank	0.000	03:01
Air Blank	0.000	03:03
Subject Sample # 2	0.138	03:04
Breath Volume	2.464	03:04
Air Blank	0.000	03:05
Control Test	0.081	03:05
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

Data Download Date/Time: 03/20/2019 23:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 02/09/2019 04:35:01

Diagnostics Check	OK	04:28
Air Blank	0.000	04:28
Control Test	0.081	04:29
Air Blank	0.000	04:29
Subject Sample # 1	0.096	04:30
Breath Volume	2.914	04:30
Air Blank	0.000	04:30
Air Blank	0.000	04:32
Subject Sample # 2	0.096	04:33
Breath Volume	3.328	04:33
Air Blank	0.000	04:34
Control Test	0.080	04:34
Air Blank	0.000	04:34
Diagnostics Check	OK	04:35

Data Download Date/Time: 03/20/2019 23:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEDE,BRENNEN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WEDE,BRENNEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: MAY

DATE/TIME OF TEST: 02/10/2019 02:39:48

Diagnostics Check	OK	02:36	
Air Blank	0.000	02:37	
Control Test	0.081	02:37	
Air Blank	0.000	02:38	
Subject Sample	# 1	02:38	Subject Test Refused
Breath Volume	0.000	02:38	
Air Blank	0.000	02:38	
Control Test	0.081	02:39	
Air Blank	0.000	02:39	
Diagnostics Check	OK	02:39	

Data Download Date/Time: 03/20/2019 23:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: ST PAUL

DATE/TIME OF TEST: 02/10/2019 03:37:37

Diagnostics Check	OK	03:31
Air Blank	0.000	03:32
Control Test	0.082	03:32
Air Blank	0.000	03:32
Subject Sample # 1	0.147	03:33
Breath Volume	2.125	03:33
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.144	03:36
Breath Volume	2.253	03:36
Air Blank	0.000	03:36
Control Test	0.081	03:37
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 03/20/2019 23:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BEHRENSHOUSER,RICHARD
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 02/10/2019 05:11:02

Diagnostics Check	OK	05:04
Air Blank	0.000	05:05
Control Test	0.081	05:05
Air Blank	0.000	05:06
Subject Sample # 1	0.146	05:06
Breath Volume	1.453	05:06
Air Blank	0.000	05:07
Air Blank	0.000	05:09
Subject Sample # 2	0.141	05:09
Breath Volume	1.835	05:09
Air Blank	0.000	05:10
Control Test	0.080	05:10
Air Blank	0.000	05:10
Diagnostics Check	OK	05:11

Data Download Date/Time: 03/20/2019 23:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, WILLIAM, T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: COTTON

DATE/TIME OF TEST: 02/11/2019 04:13:18

Diagnostics Check	OK	04:01	
Air Blank	0.000	04:02	
Control Test	0.082	04:02	
Air Blank	0.000	04:03	
Subject Sample	# 1	04:06	No Sample Provided
Breath Volume	0.000	04:06	
Air Blank	0.000	04:06	
Air Blank	0.000	04:08	
Subject Sample	# 2	04:11	No Sample Provided
Breath Volume	0.000	04:11	
Air Blank	0.000	04:12	
Control Test	0.082	04:12	
Air Blank	0.000	04:13	
Diagnostics Check	OK	04:13	

Data Download Date/Time: 03/20/2019 23:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PALAZZOLO,COURTNEY
ARRESTING OFFICER AGENCY: Gulfport PD
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: FRANZESE

DATE/TIME OF TEST: 02/12/2019 22:40:40

Diagnostics Check	OK	22:37	
Air Blank	0.000	22:37	
Control Test	0.082	22:38	
Air Blank	0.000	22:38	
Subject Sample	# 1	22:39	Subject Test Refused
Breath Volume	0.000	22:39	
Air Blank	0.000	22:39	
Control Test	0.081	22:40	
Air Blank	0.000	22:40	
Diagnostics Check	OK	22:40	

Data Download Date/Time: 03/20/2019 23:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICOTTILLI, JOSHUA, A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE, LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: SOBOLEWSKI

DATE/TIME OF TEST: 02/14/2019 04:55:26

Diagnostics Check	OK	04:50	
Air Blank	0.000	04:51	
Control Test	0.082	04:51	
Air Blank	0.000	04:52	
Subject Sample	# 1	04:54	Subject Test Refused
Breath Volume	0.000	04:54	
Air Blank	0.000	04:54	
Control Test	0.082	04:54	
Air Blank	0.000	04:55	
Diagnostics Check	OK	04:55	

Data Download Date/Time: 03/20/2019 23:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: HARWOOD

DATE/TIME OF TEST: 02/15/2019 02:47:19

Diagnostics Check	OK	02:35	
Air Blank	0.000	02:36	
Control Test	0.082	02:36	
Air Blank	0.000	02:37	
Subject Sample	# 1	02:40	No Sample Provided
Breath Volume	0.000	02:40	
Air Blank	0.000	02:40	
Air Blank	0.000	02:42	
Subject Sample	# 2	02:45	No Sample Provided
Breath Volume	0.000	02:45	
Air Blank	0.000	02:46	
Control Test	0.082	02:46	
Air Blank	0.000	02:47	
Diagnostics Check	OK	02:47	

Data Download Date/Time: 03/20/2019 23:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: ZIEGLER

DATE/TIME OF TEST: 02/15/2019 02:58:27

Diagnostics Check	OK	02:55	
Air Blank	0.000	02:56	
Control Test	0.082	02:56	
Air Blank	0.000	02:56	
Subject Sample	# 1	02:57	Subject Test Refused
Breath Volume	0.000	02:57	
Air Blank	0.000	02:57	
Control Test	0.081	02:57	
Air Blank	0.000	02:58	
Diagnostics Check	OK	02:58	

Data Download Date/Time: 03/20/2019 23:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: BIELICKI

DATE/TIME OF TEST: 02/15/2019 05:08:41

Diagnostics Check	OK	04:57	
Air Blank	0.000	04:57	
Control Test	0.082	04:58	
Air Blank	0.000	04:58	
Subject Sample	# 1	05:01	No Sample Provided
Breath Volume	0.000	05:01	
Air Blank	0.000	05:02	
Air Blank	0.000	05:04	
Subject Sample	# 2	05:07	No Sample Provided
Breath Volume	0.000	05:07	
Air Blank	0.000	05:07	
Control Test	0.082	05:08	
Air Blank	0.000	05:08	
Diagnostics Check	OK	05:08	

Data Download Date/Time: 03/20/2019 23:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FEIDT,E
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:23

SUBJECT'S LAST NAME: MCCracken

DATE/TIME OF TEST: 02/15/2019 19:47:00

Diagnostics Check	OK	19:44	
Air Blank	0.000	19:44	
Control Test	0.082	19:44	
Air Blank	0.000	19:45	
Subject Sample	# 1	19:45	Subject Test Refused
Breath Volume	0.000	19:45	
Air Blank	0.000	19:46	
Control Test	0.082	19:46	
Air Blank	0.000	19:46	
Diagnostics Check	OK	19:47	

Data Download Date/Time: 03/20/2019 23:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: MARTIN RAMIREZ

DATE/TIME OF TEST: 02/15/2019 21:03:56

Diagnostics Check	OK	20:57
Air Blank	0.000	20:58
Control Test	0.081	20:58
Air Blank	0.000	20:58
Subject Sample # 1	0.194	20:59
Breath Volume	3.058	20:59
Air Blank	0.000	21:00
Air Blank	0.000	21:01
Subject Sample # 2	0.190	21:02
Breath Volume	3.496	21:02
Air Blank	0.000	21:03
Control Test	0.080	21:03
Air Blank	0.000	21:03
Diagnostics Check	OK	21:03

Data Download Date/Time: 03/20/2019 23:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, WILLIAM, T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: NGUYEN

DATE/TIME OF TEST: 02/16/2019 02:43:11

Diagnostics Check	OK	02:36
Air Blank	0.000	02:37
Control Test	0.082	02:37
Air Blank	0.000	02:38
Subject Sample # 1	0.145	02:38
Breath Volume	3.222	02:38
Air Blank	0.000	02:39
Air Blank	0.000	02:41
Subject Sample # 2	0.148	02:41
Breath Volume	3.210	02:41
Air Blank	0.000	02:42
Control Test	0.081	02:42
Air Blank	0.000	02:43
Diagnostics Check	OK	02:43

Data Download Date/Time: 03/20/2019 23:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHCOM, TIFFANY, J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ASHCOM, TIFFANY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: ARCE HERRERA

DATE/TIME OF TEST: 02/16/2019 03:10:29

Diagnostics Check	OK	03:04
Air Blank	0.000	03:05
Control Test	0.081	03:05
Air Blank	0.000	03:05
Subject Sample # 1	0.000	03:06
Breath Volume	1.480	03:06
Air Blank	0.000	03:06
Air Blank	0.000	03:08
Subject Sample # 2	0.000	03:09
Breath Volume	1.898	03:09
Air Blank	0.000	03:09
Control Test	0.081	03:09
Air Blank	0.000	03:10
Diagnostics Check	OK	03:10

Data Download Date/Time: 03/20/2019 23:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, WILLIAM, T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BEHRENSHOUSER, RICHARD
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 02/16/2019 03:43:19

Diagnostics Check	OK	03:36
Air Blank	0.000	03:36
Control Test	0.081	03:37
Air Blank	0.000	03:37
Subject Sample # 1	0.127	03:39
Breath Volume	2.265	03:39
Air Blank	0.000	03:39
Air Blank	0.000	03:41
Subject Sample # 2	0.122	03:41
Breath Volume	1.371	03:41
Air Blank	0.000	03:42
Control Test	0.081	03:42
Air Blank	0.000	03:43
Diagnostics Check	OK	03:43

Data Download Date/Time: 03/20/2019 23:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHCOM, TIFFANY, J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ASHCOM, TIFFANY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:22

SUBJECT'S LAST NAME: KUDRYASHOV

DATE/TIME OF TEST: 02/16/2019 18:48:42

Diagnostics Check	OK	18:42
Air Blank	0.000	18:42
Control Test	0.082	18:43
Air Blank	0.000	18:43
Subject Sample # 1	0.178	18:44
Breath Volume	1.625	18:44
Air Blank	0.000	18:44
Air Blank	0.000	18:46
Subject Sample # 2	0.196	18:47
Breath Volume	2.917	18:47
Air Blank	0.000	18:47
Control Test	0.081	18:48
Air Blank	0.000	18:48
Diagnostics Check	OK	18:48

Data Download Date/Time: 03/20/2019 23:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BEHRENSHOUSER,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 02/17/2019 03:08:03

Diagnostics Check	OK	03:05	
Air Blank	0.000	03:05	
Control Test	0.082	03:06	
Air Blank	0.000	03:06	
Subject Sample	# 1	03:06	Subject Test Refused
Breath Volume	0.000	03:06	
Air Blank	0.000	03:07	
Control Test	0.081	03:07	
Air Blank	0.000	03:07	
Diagnostics Check	OK	03:08	

Data Download Date/Time: 03/20/2019 23:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GRILL,S
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: VENABLE

DATE/TIME OF TEST: 02/17/2019 03:23:33

Diagnostics Check	OK	03:20	
Air Blank	0.000	03:21	
Control Test	0.082	03:21	
Air Blank	0.000	03:22	
Subject Sample	# 1	03:22	Subject Test Refused
Breath Volume	0.000	03:22	
Air Blank	0.000	03:22	
Control Test	0.082	03:22	
Air Blank	0.000	03:23	
Diagnostics Check	OK	03:23	

Data Download Date/Time: 03/20/2019 23:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AKINS,G
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: CASTRO

DATE/TIME OF TEST: 02/17/2019 04:12:28

Diagnostics Check	OK	04:09	
Air Blank	0.000	04:10	
Control Test	0.082	04:10	
Air Blank	0.000	04:10	
Subject Sample	# 1	04:11	Subject Test Refused
Breath Volume	0.000	04:11	
Air Blank	0.000	04:11	
Control Test	0.082	04:11	
Air Blank	0.000	04:12	
Diagnostics Check	OK	04:12	

Data Download Date/Time: 03/20/2019 23:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,KYLE,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CAHOON,MARK
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: HANDY

DATE/TIME OF TEST: 02/17/2019 05:19:57

Diagnostics Check	OK	05:17	
Air Blank	0.000	05:17	
Control Test	0.082	05:17	
Air Blank	0.000	05:18	
Subject Sample	# 1	05:18	Subject Test Refused
Breath Volume	0.000	05:18	
Air Blank	0.000	05:19	
Control Test	0.082	05:19	
Air Blank	0.000	05:19	
Diagnostics Check	OK	05:19	

Data Download Date/Time: 03/20/2019 23:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADA,RAYMOND,O
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ADA,RAYMOND
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 08:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:45

SUBJECT'S LAST NAME: JEUP

DATE/TIME OF TEST: 02/17/2019 10:15:20

Diagnostics Check	OK	10:05
Air Blank	0.000	10:06
Control Test	0.082	10:06
Air Blank	0.000	10:07
Subject Sample # 1	0.286	10:07
Breath Volume	2.769	10:07
Air Blank	0.000	10:08
Air Blank	0.000	10:10
Subject Sample # 2	0.255	10:10
Breath Volume	1.859	10:10
Air Blank	0.000	10:11
Air Blank	0.000	10:13
Subject Sample # 3	0.285	10:13
Breath Volume	2.972	10:13
Air Blank	0.000	10:14
Control Test	0.080	10:14
Air Blank	0.000	10:15
Diagnostics Check	OK	10:15

Data Download Date/Time: 03/20/2019 23:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,KYLE,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GRIFFIN,KYLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:53

SUBJECT'S LAST NAME: DENNY

DATE/TIME OF TEST: 02/17/2019 20:19:08

Diagnostics Check	OK	20:15	
Air Blank	0.000	20:16	
Control Test	0.083	20:16	
Air Blank	0.000	20:17	
Subject Sample	# 1	20:17	Subject Test Refused
Breath Volume	0.000	20:17	
Air Blank	0.000	20:18	
Control Test	0.081	20:18	
Air Blank	0.000	20:19	
Diagnostics Check	OK	20:19	

Data Download Date/Time: 03/20/2019 23:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, WILLIAM, T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:22

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 02/20/2019 05:51:14

Diagnostics Check	OK	05:44
Air Blank	0.000	05:44
Control Test	0.082	05:44
Air Blank	0.000	05:45
Subject Sample # 1	0.059	05:46
Breath Volume	1.617	05:46
Air Blank	0.000	05:47
Air Blank	0.000	05:49
Subject Sample # 2	0.059	05:49
Breath Volume	1.347	05:49
Air Blank	0.000	05:50
Control Test	0.081	05:50
Air Blank	0.000	05:51
Diagnostics Check	OK	05:51

Data Download Date/Time: 03/20/2019 23:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBY,LARRY,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: JACOBY,LARRY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: CONNOR

DATE/TIME OF TEST: 02/20/2019 23:55:25

Diagnostics Check	OK	23:49
Air Blank	0.000	23:49
Control Test	0.082	23:49
Air Blank	0.000	23:50
Subject Sample # 1	0.130	23:50
Breath Volume	1.179	23:50
Air Blank	0.000	23:51
Air Blank	0.000	23:53
Subject Sample # 2	0.121	23:53
Breath Volume	1.675	23:53
Air Blank	0.000	23:54
Control Test	0.081	23:54
Air Blank	0.000	23:55
Diagnostics Check	OK	23:55

Data Download Date/Time: 03/20/2019 23:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LEIB,M
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: FANNING

DATE/TIME OF TEST: 02/21/2019 00:44:32

Diagnostics Check	OK	00:37
Air Blank	0.000	00:37
Control Test	0.082	00:38
Air Blank	0.000	00:38
Subject Sample # 1	0.097	00:39
Breath Volume	1.769	00:39
Air Blank	0.000	00:40
Air Blank	0.000	00:41
Subject Sample # 2	0.097	00:42
Breath Volume	1.671	00:42
Air Blank	0.000	00:43
Control Test	0.081	00:43
Air Blank	0.000	00:44
Diagnostics Check	OK	00:44

Data Download Date/Time: 03/20/2019 23:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEDE,BRENNEN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WEDE,BRENNEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: HALLAM

DATE/TIME OF TEST: 02/22/2019 02:38:42

Diagnostics Check	OK	02:31
Air Blank	0.000	02:32
Control Test	0.082	02:32
Air Blank	0.000	02:32
Subject Sample # 1	0.107	02:34
Breath Volume	1.453	02:34
Air Blank	0.000	02:35
Air Blank	0.000	02:36
Subject Sample # 2	0.117	02:37
Breath Volume	1.242	02:37
Air Blank	0.000	02:37
Control Test	0.081	02:38
Air Blank	0.000	02:38
Diagnostics Check	OK	02:38

Data Download Date/Time: 03/20/2019 23:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RANISZEWSKI,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: JENKS

DATE/TIME OF TEST: 02/22/2019 03:43:09

Diagnostics Check	OK	03:35
Air Blank	0.000	03:36
Control Test	0.082	03:36
Air Blank	0.000	03:36
Subject Sample # 1	0.234	03:38
Breath Volume	1.324	03:38
Air Blank	0.000	03:38
Air Blank	0.000	03:40
Subject Sample # 2	0.244	03:41
Breath Volume	2.070	03:41
Air Blank	0.000	03:42
Control Test	0.081	03:42
Air Blank	0.000	03:43
Diagnostics Check	OK	03:43

Data Download Date/Time: 03/20/2019 23:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEDE,BRENNEN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GARDNER,DANIEL
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:18

SUBJECT'S LAST NAME: YOST

DATE/TIME OF TEST: 02/22/2019 05:54:36

Diagnostics Check	OK	05:38	
Air Blank	0.000	05:39	
Control Test	0.081	05:39	
Air Blank	0.000	05:40	
Subject Sample	# 1 0.117	05:43	Volume Not Met
Breath Volume	0.230	05:43	
Air Blank	0.000	05:43	
Air Blank	0.000	05:45	
Subject Sample	# 2 0.117	05:47	
Breath Volume	1.148	05:47	
Air Blank	0.000	05:48	
Air Blank	0.000	05:50	
Subject Sample	# 3	05:53	No Sample Provided
Breath Volume	0.000	05:53	
Air Blank	0.000	05:53	
Control Test	0.080	05:54	
Air Blank	0.000	05:54	
Diagnostics Check	OK	05:54	

Data Download Date/Time: 03/20/2019 23:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,L
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: CIACCIO

DATE/TIME OF TEST: 02/23/2019 00:15:00

Diagnostics Check	OK	00:07
Air Blank	0.000	00:07
Control Test	0.082	00:08
Air Blank	0.000	00:08
Subject Sample # 1	0.043	00:09
Breath Volume	1.937	00:09
Air Blank	0.000	00:10
Air Blank	0.000	00:12
Subject Sample # 2	0.040	00:13
Breath Volume	1.535	00:13
Air Blank	0.000	00:14
Control Test	0.081	00:14
Air Blank	0.000	00:14
Diagnostics Check	OK	00:15

Data Download Date/Time: 03/20/2019 23:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RANISZEWSKI,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: JACOBS

DATE/TIME OF TEST: 02/23/2019 04:01:48

Diagnostics Check	OK	03:57	
Air Blank	0.000	03:58	
Control Test	0.081	03:58	
Air Blank	0.000	03:59	
Subject Sample	# 1	04:00	Subject Test Refused
Breath Volume	0.000	04:00	
Air Blank	0.000	04:00	
Control Test	0.081	04:01	
Air Blank	0.000	04:01	
Diagnostics Check	OK	04:01	

Data Download Date/Time: 03/20/2019 23:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEDE,BRENNEN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WEDE,BRENNEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: LENTZ

DATE/TIME OF TEST: 02/23/2019 04:32:07

Diagnostics Check	OK	04:27	
Air Blank	0.000	04:27	
Control Test	0.081	04:27	
Air Blank	0.000	04:28	
Subject Sample	# 1	04:30	Subject Test Refused
Breath Volume	0.000	04:30	
Air Blank	0.000	04:31	
Control Test	0.081	04:31	
Air Blank	0.000	04:32	
Diagnostics Check	OK	04:32	

Data Download Date/Time: 03/20/2019 23:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEDE,BRENNEN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WEDE,BRENNEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: LENTZ

DATE/TIME OF TEST: 02/23/2019 05:15:50

Diagnostics Check	OK	05:03	
Air Blank	0.000	05:04	
Control Test	0.081	05:04	
Air Blank	0.000	05:04	
Subject Sample	# 1 0.130	05:08	Volume Not Met
Breath Volume	0.289	05:08	
Air Blank	0.000	05:08	
Air Blank	0.000	05:10	
Subject Sample	# 2 0.148	05:11	
Breath Volume	2.375	05:11	
Air Blank	0.000	05:12	
Air Blank	0.000	05:14	
Subject Sample	# 3 0.151	05:14	
Breath Volume	2.277	05:14	
Air Blank	0.000	05:14	
Control Test	0.080	05:15	
Air Blank	0.000	05:15	
Diagnostics Check	OK	05:15	

Data Download Date/Time: 03/20/2019 23:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,L
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: FAIRBANKS

DATE/TIME OF TEST: 02/24/2019 01:55:29

Diagnostics Check	OK	01:48
Air Blank	0.000	01:49
Control Test	0.082	01:49
Air Blank	0.000	01:50
Subject Sample # 1	0.118	01:50
Breath Volume	3.015	01:50
Air Blank	0.000	01:51
Air Blank	0.000	01:53
Subject Sample # 2	0.121	01:53
Breath Volume	2.792	01:53
Air Blank	0.000	01:54
Control Test	0.081	01:54
Air Blank	0.000	01:55
Diagnostics Check	OK	01:55

Data Download Date/Time: 03/20/2019 23:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RANISZEWSKI,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 02/24/2019 02:36:41

Diagnostics Check	OK	02:22	
Air Blank	0.000	02:23	
Control Test	0.081	02:23	
Air Blank	0.000	02:24	
Subject Sample	# 1 0.128	02:27	Slope Not Level
Breath Volume	1.121	02:27	
Air Blank	0.000	02:27	
Air Blank	0.000	02:29	
Subject Sample	# 2 0.124	02:31	
Breath Volume	1.964	02:31	
Air Blank	0.000	02:32	
Air Blank	0.000	02:34	
Subject Sample	# 3 0.127	02:35	
Breath Volume	2.035	02:35	
Air Blank	0.000	02:35	
Control Test	0.081	02:36	
Air Blank	0.000	02:36	
Diagnostics Check	OK	02:36	

Data Download Date/Time: 03/20/2019 23:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LAMBORGHINI,P
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 02/24/2019 04:47:37

Diagnostics Check	OK	04:40
Air Blank	0.000	04:40
Control Test	0.082	04:41
Air Blank	0.000	04:41
Subject Sample # 1	0.157	04:42
Breath Volume	3.683	04:42
Air Blank	0.000	04:43
Air Blank	0.000	04:45
Subject Sample # 2	0.157	04:46
Breath Volume	3.468	04:46
Air Blank	0.000	04:46
Control Test	0.080	04:47
Air Blank	0.000	04:47
Diagnostics Check	OK	04:47

Data Download Date/Time: 03/20/2019 23:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEDE,BRENNEN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: BEADLE

DATE/TIME OF TEST: 02/24/2019 23:32:51

Diagnostics Check	OK	23:26
Air Blank	0.000	23:27
Control Test	0.082	23:27
Air Blank	0.000	23:27
Subject Sample # 1	0.056	23:28
Breath Volume	1.457	23:28
Air Blank	0.000	23:28
Air Blank	0.000	23:30
Subject Sample # 2	0.060	23:31
Breath Volume	1.457	23:31
Air Blank	0.000	23:31
Control Test	0.081	23:32
Air Blank	0.000	23:32
Diagnostics Check	OK	23:32

Data Download Date/Time: 03/20/2019 23:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICOTTILLI, JOSHUA, A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CASTELEIRO, ANTHONY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: GJOKA

DATE/TIME OF TEST: 02/25/2019 02:10:35

Diagnostics Check	OK	02:03
Air Blank	0.000	02:03
Control Test	0.082	02:03
Air Blank	0.000	02:04
Subject Sample # 1	0.193	02:05
Breath Volume	1.402	02:05
Air Blank	0.000	02:06
Air Blank	0.000	02:08
Subject Sample # 2	0.191	02:09
Breath Volume	1.433	02:09
Air Blank	0.000	02:09
Control Test	0.081	02:09
Air Blank	0.000	02:10
Diagnostics Check	OK	02:10

Data Download Date/Time: 03/20/2019 23:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEDE,BRENNEN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WEDE,BRENNEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: ROBLES

DATE/TIME OF TEST: 02/25/2019 03:34:45

Diagnostics Check	OK	03:27
Air Blank	0.000	03:28
Control Test	0.082	03:28
Air Blank	0.000	03:29
Subject Sample # 1	0.175	03:30
Breath Volume	1.437	03:30
Air Blank	0.000	03:30
Air Blank	0.000	03:32
Subject Sample # 2	0.175	03:33
Breath Volume	1.500	03:33
Air Blank	0.000	03:33
Control Test	0.081	03:34
Air Blank	0.000	03:34
Diagnostics Check	OK	03:34

Data Download Date/Time: 03/20/2019 23:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICOTTILLI, JOSHUA, A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI, JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: STURGIS

DATE/TIME OF TEST: 02/25/2019 04:37:39

Diagnostics Check	OK	04:28
Air Blank	0.000	04:29
Control Test	0.082	04:29
Air Blank	0.000	04:30
Subject Sample # 1	0.165	04:32
Breath Volume	1.445	04:32
Air Blank	0.000	04:32
Air Blank	0.000	04:34
Subject Sample # 2	0.170	04:36
Breath Volume	1.722	04:36
Air Blank	0.000	04:36
Control Test	0.081	04:37
Air Blank	0.000	04:37
Diagnostics Check	OK	04:37

Data Download Date/Time: 03/20/2019 23:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUDTKE,CHRISTOPHER,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LUDTKE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: PRATT

DATE/TIME OF TEST: 02/26/2019 03:29:09

Diagnostics Check	OK	03:20
Air Blank	0.000	03:21
Control Test	0.082	03:21
Air Blank	0.000	03:22
Subject Sample # 1	0.147	03:24
Breath Volume	3.335	03:24
Air Blank	0.000	03:25
Air Blank	0.000	03:26
Subject Sample # 2	0.141	03:27
Breath Volume	2.085	03:27
Air Blank	0.000	03:28
Control Test	0.082	03:28
Air Blank	0.000	03:29
Diagnostics Check	OK	03:29

Data Download Date/Time: 03/20/2019 23:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHCOM, TIFFANY, J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZSIDO, DANIELLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: MONTENARE

DATE/TIME OF TEST: 02/27/2019 01:37:28

Diagnostics Check	OK	01:34	
Air Blank	0.000	01:35	
Control Test	0.082	01:35	
Air Blank	0.000	01:35	
Subject Sample	# 1	01:36	Subject Test Refused
Breath Volume	0.000	01:36	
Air Blank	0.000	01:36	
Control Test	0.082	01:36	
Air Blank	0.000	01:37	
Diagnostics Check	OK	01:37	

Data Download Date/Time: 03/20/2019 23:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEDE,BRENNEN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WEDE,BRENNEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: GIBBONS

DATE/TIME OF TEST: 02/28/2019 02:11:12

Diagnostics Check	OK	02:08	
Air Blank	0.000	02:08	
Control Test	0.082	02:09	
Air Blank	0.000	02:09	
Subject Sample	# 1	02:09	Subject Test Refused
Breath Volume	0.000	02:09	
Air Blank	0.000	02:10	
Control Test	0.081	02:10	
Air Blank	0.000	02:11	
Diagnostics Check	OK	02:11	

Data Download Date/Time: 03/20/2019 23:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: SEUNBANE

DATE/TIME OF TEST: 02/28/2019 02:49:44

Diagnostics Check	OK	02:43
Air Blank	0.000	02:43
Control Test	0.082	02:44
Air Blank	0.000	02:44
Subject Sample # 1	0.152	02:45
Breath Volume	1.539	02:45
Air Blank	0.000	02:45
Air Blank	0.000	02:47
Subject Sample # 2	0.163	02:48
Breath Volume	1.531	02:48
Air Blank	0.000	02:48
Control Test	0.081	02:49
Air Blank	0.000	02:49
Diagnostics Check	OK	02:49

Data Download Date/Time: 03/20/2019 23:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RANISZEWSKI,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: GARRETT

DATE/TIME OF TEST: 02/28/2019 03:32:46

Diagnostics Check	OK	03:21	
Air Blank	0.000	03:21	
Control Test	0.082	03:21	
Air Blank	0.000	03:22	
Subject Sample	# 1 0.184	03:25	Volume Not Met
Breath Volume	0.519	03:25	
Air Blank	0.000	03:26	
Air Blank	0.000	03:28	
Subject Sample	# 2 0.208	03:31	Volume Not Met
Breath Volume	0.093	03:31	
Air Blank	0.000	03:31	
Control Test	0.081	03:32	
Air Blank	0.000	03:32	
Diagnostics Check	OK	03:32	

Data Download Date/Time: 03/20/2019 23:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RANISZEWSKI,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: GARRETT

DATE/TIME OF TEST: 02/28/2019 03:43:49

Diagnostics Check	OK	03:34
Air Blank	0.000	03:35
Control Test	0.080	03:35
Air Blank	0.000	03:36
Subject Sample # 1	0.219	03:37
Breath Volume	2.539	03:37
Air Blank	0.000	03:38
Air Blank	0.000	03:40
Subject Sample # 2	0.213	03:42
Breath Volume	2.230	03:42
Air Blank	0.000	03:42
Control Test	0.079	03:43
Air Blank	0.000	03:43
Diagnostics Check	OK	03:43

Data Download Date/Time: 03/20/2019 23:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REEVES,MARK,O
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PICARD,R
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:08

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 02/28/2019 17:31:21

Diagnostics Check	OK	17:28	
Air Blank	0.000	17:28	
Control Test	0.082	17:29	
Air Blank	0.000	17:29	
Subject Sample	# 1	17:30	Subject Test Refused
Breath Volume	0.000	17:30	
Air Blank	0.000	17:30	
Control Test	0.082	17:30	
Air Blank	0.000	17:31	
Diagnostics Check	OK	17:31	

Data Download Date/Time: 03/20/2019 23:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RANISZEWSKI,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: MCKEE

DATE/TIME OF TEST: 02/28/2019 23:00:20

Diagnostics Check	OK	22:57	
Air Blank	0.000	22:57	
Control Test	0.082	22:58	
Air Blank	0.000	22:58	
Subject Sample	# 1	22:58	Subject Test Refused
Breath Volume	0.000	22:58	
Air Blank	0.000	22:59	
Control Test	0.082	22:59	
Air Blank	0.000	23:00	
Diagnostics Check	OK	23:00	

Data Download Date/Time: 03/20/2019 23:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUGENT,DAVID,A
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NUGENT,DAVID
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:31

SUBJECT'S LAST NAME: GRINBERG

DATE/TIME OF TEST: 02/01/2019 18:59:36

Diagnostics Check	OK	18:53
Air Blank	0.000	18:53
Control Test	0.080	18:54
Air Blank	0.000	18:54
Subject Sample # 1	0.000	18:55
Breath Volume	1.882	18:55
Air Blank	0.000	18:55
Air Blank	0.000	18:57
Subject Sample # 2	0.000	18:58
Breath Volume	2.117	18:58
Air Blank	0.000	18:58
Control Test	0.080	18:58
Air Blank	0.000	18:59
Diagnostics Check	OK	18:59

Data Download Date/Time: 02/20/2019 20:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEGERSMITH,DANIEL,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,DANIEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: BIANCO

DATE/TIME OF TEST: 02/01/2019 21:58:19

Diagnostics Check	OK	21:51
Air Blank	0.000	21:51
Control Test	0.079	21:52
Air Blank	0.000	21:52
Subject Sample # 1	0.261	21:53
Breath Volume	2.316	21:53
Air Blank	0.000	21:54
Air Blank	0.000	21:55
Subject Sample # 2	0.255	21:56
Breath Volume	2.503	21:56
Air Blank	0.000	21:57
Control Test	0.079	21:57
Air Blank	0.000	21:58
Diagnostics Check	OK	21:58

Data Download Date/Time: 02/20/2019 20:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HURT,JONATHAN,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,DANIEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: CERNAN

DATE/TIME OF TEST: 02/03/2019 01:44:30

Diagnostics Check	OK	01:37
Air Blank	0.000	01:37
Control Test	0.079	01:37
Air Blank	0.000	01:38
Subject Sample # 1	0.071	01:39
Breath Volume	3.109	01:39
Air Blank	0.000	01:40
Air Blank	0.000	01:42
Subject Sample # 2	0.071	01:42
Breath Volume	2.921	01:42
Air Blank	0.000	01:43
Control Test	0.079	01:43
Air Blank	0.000	01:44
Diagnostics Check	OK	01:44

Data Download Date/Time: 02/20/2019 20:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HURT,JONATHAN,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: HURT,JONATHAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: OWEN JR

DATE/TIME OF TEST: 02/03/2019 02:02:28

Diagnostics Check	OK	01:55
Air Blank	0.000	01:56
Control Test	0.079	01:56
Air Blank	0.000	01:56
Subject Sample # 1	0.000	01:57
Breath Volume	1.257	01:57
Air Blank	0.000	01:58
Air Blank	0.000	02:00
Subject Sample # 2	0.000	02:01
Breath Volume	1.613	02:01
Air Blank	0.000	02:01
Control Test	0.084	02:01
Air Blank	0.000	02:02
Diagnostics Check	OK	02:02

Data Download Date/Time: 02/20/2019 20:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HURT,JONATHAN,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: HURT,JONATHAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: BARBOZA TORRESS

DATE/TIME OF TEST: 02/09/2019 19:33:35

Diagnostics Check	OK	19:26
Air Blank	0.000	19:26
Control Test	0.079	19:27
Air Blank	0.000	19:27
Subject Sample # 1	0.087	19:28
Breath Volume	3.578	19:28
Air Blank	0.000	19:29
Air Blank	0.000	19:31
Subject Sample # 2	0.090	19:32
Breath Volume	3.300	19:32
Air Blank	0.000	19:32
Control Test	0.079	19:32
Air Blank	0.000	19:33
Diagnostics Check	OK	19:33

Data Download Date/Time: 02/20/2019 20:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUIS,OFC. J,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: BUIS,OFC. J
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 02/15/2019 05:11:18

Diagnostics Check	OK	05:04
Air Blank	0.000	05:05
Control Test	0.079	05:05
Air Blank	0.000	05:05
Subject Sample # 1	0.178	05:06
Breath Volume	2.437	05:06
Air Blank	0.000	05:07
Air Blank	0.000	05:09
Subject Sample # 2	0.175	05:09
Breath Volume	2.023	05:09
Air Blank	0.000	05:10
Control Test	0.078	05:10
Air Blank	0.000	05:11
Diagnostics Check	OK	05:11

Data Download Date/Time: 02/20/2019 20:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEGERSMITH,DANIEL,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,DANIEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: KERNS

DATE/TIME OF TEST: 02/15/2019 18:43:56

Diagnostics Check	OK	18:37
Air Blank	0.000	18:37
Control Test	0.079	18:38
Air Blank	0.000	18:38
Subject Sample # 1	0.000	18:39
Breath Volume	4.054	18:39
Air Blank	0.000	18:39
Air Blank	0.000	18:41
Subject Sample # 2	0.000	18:42
Breath Volume	2.773	18:42
Air Blank	0.000	18:43
Control Test	0.079	18:43
Air Blank	0.000	18:43
Diagnostics Check	OK	18:43

Data Download Date/Time: 02/20/2019 20:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HURT,JONATHAN,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,DANIEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 02/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: HURST

DATE/TIME OF TEST: 02/21/2019 01:39:32

Diagnostics Check	OK	01:30
Air Blank	0.000	01:31
Control Test	0.080	01:31
Air Blank	0.000	01:32
Subject Sample # 1	0.143	01:33
Breath Volume	1.789	01:33
Air Blank	0.000	01:34
Air Blank	0.000	01:36
Subject Sample # 2	0.141	01:37
Breath Volume	2.093	01:37
Air Blank	0.000	01:38
Control Test	0.079	01:38
Air Blank	0.000	01:39
Diagnostics Check	OK	01:39

Data Download Date/Time: 03/13/2019 15:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HURT,JONATHAN,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: HURT,JONATHAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 02/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: SIEV

DATE/TIME OF TEST: 02/22/2019 03:52:55

Diagnostics Check	OK	03:45
Air Blank	0.000	03:46
Control Test	0.079	03:46
Air Blank	0.000	03:47
Subject Sample # 1	0.169	03:48
Breath Volume	2.917	03:48
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 2	0.168	03:51
Breath Volume	3.125	03:51
Air Blank	0.000	03:52
Control Test	0.079	03:52
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time: 03/13/2019 15:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HURT,JONATHAN,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: HURT,JONATHAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 02/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: VALOIS

DATE/TIME OF TEST: 02/23/2019 01:08:18

Diagnostics Check	OK	01:00
Air Blank	0.000	01:00
Control Test	0.079	01:01
Air Blank	0.000	01:01
Subject Sample # 1	0.315	01:03
Breath Volume	1.164	01:03
Air Blank	0.000	01:03
Air Blank	0.000	01:05
Subject Sample # 2	0.303	01:06
Breath Volume	1.429	01:06
Air Blank	0.000	01:07
Control Test	0.079	01:07
Air Blank	0.000	01:08
Diagnostics Check	OK	01:08

Data Download Date/Time: 03/13/2019 15:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUGENT,DAVID,A
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NUGENT,DAVID
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 02/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: TICE

DATE/TIME OF TEST: 02/24/2019 00:37:24

Diagnostics Check	OK	00:25	
Air Blank	0.000	00:26	
Control Test	0.080	00:26	
Air Blank	0.000	00:27	
Subject Sample	# 1	00:30	No Sample Provided
Breath Volume	0.000	00:30	
Air Blank	0.000	00:30	
Air Blank	0.000	00:32	
Subject Sample	# 2	00:35	No Sample Provided
Breath Volume	0.000	00:35	
Air Blank	0.000	00:36	
Control Test	0.080	00:36	
Air Blank	0.000	00:37	
Diagnostics Check	OK	00:37	

Data Download Date/Time: 03/13/2019 15:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUGENT,DAVID,A
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NUGENT,DAVID
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 02/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: WREN

DATE/TIME OF TEST: 02/28/2019 01:42:50

Diagnostics Check	OK	01:35
Air Blank	0.000	01:35
Control Test	0.079	01:35
Air Blank	0.000	01:36
Subject Sample # 1	0.177	01:37
Breath Volume	2.105	01:37
Air Blank	0.000	01:38
Air Blank	0.000	01:39
Subject Sample # 2	0.166	01:41
Breath Volume	2.488	01:41
Air Blank	0.000	01:41
Control Test	0.080	01:42
Air Blank	0.000	01:42
Diagnostics Check	OK	01:42

Data Download Date/Time: 03/13/2019 15:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HURT,JONATHAN,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,DANIEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 02/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: MANN

DATE/TIME OF TEST: 02/28/2019 01:52:39

Diagnostics Check	OK	01:45
Air Blank	0.000	01:45
Control Test	0.079	01:46
Air Blank	0.000	01:46
Subject Sample # 1	0.142	01:47
Breath Volume	2.542	01:47
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample # 2	0.140	01:50
Breath Volume	2.156	01:50
Air Blank	0.000	01:51
Control Test	0.077	01:52
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

Data Download Date/Time: 03/13/2019 15:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001073

INSTRUMENT REGISTERED TO: DADE CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AIKEN,DURRIEL,R
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: AIKEN,DURRIEL
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A3
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: VELEZ-MORALES

DATE/TIME OF TEST: 02/10/2019 04:32:32

Diagnostics Check	OK	04:26
Air Blank	0.000	04:26
Control Test	0.078	04:27
Air Blank	0.000	04:27
Subject Sample # 1	0.071	04:28
Breath Volume	3.382	04:28
Air Blank	0.000	04:28
Air Blank	0.000	04:30
Subject Sample # 2	0.074	04:31
Breath Volume	3.613	04:31
Air Blank	0.000	04:31
Control Test	0.080	04:31
Air Blank	0.000	04:32
Diagnostics Check	OK	04:32

Data Download Date/Time: 02/13/2019 17:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCARTHY,KEITH,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: GARMON,T
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: LARSON

DATE/TIME OF TEST: 02/05/2019 03:05:59

Diagnostics Check	OK	02:59
Air Blank	0.000	02:59
Control Test	0.077	03:00
Air Blank	0.000	03:00
Subject Sample # 1	0.284	03:01
Breath Volume	1.500	03:01
Air Blank	0.000	03:02
Air Blank	0.000	03:03
Subject Sample # 2	0.287	03:04
Breath Volume	2.300	03:04
Air Blank	0.000	03:05
Control Test	0.076	03:05
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

Data Download Date/Time: 02/19/2019 16:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,BARR,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: NIXON,BARRY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: GULLEY

DATE/TIME OF TEST: 02/08/2019 03:03:08

Diagnostics Check	OK	02:56
Air Blank	0.000	02:57
Control Test	0.077	02:57
Air Blank	0.000	02:58
Subject Sample # 1	0.129	02:58
Breath Volume	1.710	02:58
Air Blank	0.000	02:59
Air Blank	0.000	03:01
Subject Sample # 2	0.124	03:01
Breath Volume	2.074	03:01
Air Blank	0.000	03:02
Control Test	0.075	03:02
Air Blank	0.000	03:03
Diagnostics Check	OK	03:03

Data Download Date/Time: 02/19/2019 16:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUSTOS,ALEJANDRO,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BUSTOS,ALEJANDRO
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: OWEN

DATE/TIME OF TEST: 02/26/2019 02:45:27

Diagnostics Check	OK	02:38
Air Blank	0.000	02:39
Control Test	0.077	02:39
Air Blank	0.000	02:40
Subject Sample # 1	0.180	02:40
Breath Volume	2.511	02:40
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.177	02:43
Breath Volume	2.968	02:43
Air Blank	0.000	02:44
Control Test	0.075	02:44
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 03/29/2019 22:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001075

INSTRUMENT REGISTERED TO: FHP TROOP E

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BROWN,TIMOTHY,J

BREATH TEST OPERATOR AGENCY: FHP Troop E

ARRESTING OFFICER: BROWN,TIMOTHY

ARRESTING OFFICER AGENCY: FHP Troop E

LAST AGENCY INSPECTION: 01/02/2019

Date or Time Changed: Yes

DRY GAS STANDARD LOT #: 712031

DRY GAS EXPIRY DATE: 02/21/2020

VIOLATION CODE: DUI

ARREST TIME: 01:23

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: MORA

DATE/TIME OF TEST: 02/03/2019 02:19:50

Diagnostics Check	OK	02:13
Air Blank	0.000	02:13
Control Test	0.081	02:14
Air Blank	0.000	02:14
Subject Sample # 1	0.218	02:15
Breath Volume	2.445	02:15
Air Blank	0.000	02:16
Air Blank	0.000	02:17
Subject Sample # 2	0.229	02:18
Breath Volume	2.382	02:18
Air Blank	0.000	02:18
Control Test	0.080	02:19
Air Blank	0.000	02:19
Diagnostics Check	OK	02:19

Data Download Date/Time: 02/27/2019 09:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001075

INSTRUMENT REGISTERED TO: FHP TROOP E

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: NANAN,KEIL,

BREATH TEST OPERATOR AGENCY: FHP Troop E

ARRESTING OFFICER: NANAN,KEIL

ARRESTING OFFICER AGENCY: FHP Troop E

LAST AGENCY INSPECTION: 01/02/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 712031

DRY GAS EXPIRY DATE: 02/21/2020

VIOLATION CODE: DUI

ARREST TIME: 17:55

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 02/07/2019 19:10:02

Diagnostics Check	OK	19:07	
Air Blank	0.000	19:07	
Control Test	0.080	19:07	
Air Blank	0.000	19:08	
Subject Sample	# 1	19:08	Subject Test Refused
Breath Volume	0.000	19:08	
Air Blank	0.000	19:09	
Control Test	0.081	19:09	
Air Blank	0.000	19:09	
Diagnostics Check	OK	19:10	

Data Download Date/Time: 02/27/2019 09:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001075

INSTRUMENT REGISTERED TO: FHP TROOP E

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BROWN,TIMOTHY,J

BREATH TEST OPERATOR AGENCY: FHP Troop E

ARRESTING OFFICER: BROWN,TIMOTHY

ARRESTING OFFICER AGENCY: FHP Troop E

LAST AGENCY INSPECTION: 01/02/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 712031

DRY GAS EXPIRY DATE: 02/21/2020

VIOLATION CODE: DUI

ARREST TIME: 03:35

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: MASIS FLORES

DATE/TIME OF TEST: 02/08/2019 05:06:15

Diagnostics Check	OK	04:59
Air Blank	0.000	05:00
Control Test	0.081	05:00
Air Blank	0.000	05:01
Subject Sample # 1	0.189	05:01
Breath Volume	1.269	05:01
Air Blank	0.000	05:02
Air Blank	0.000	05:04
Subject Sample # 2	0.203	05:04
Breath Volume	2.343	05:04
Air Blank	0.000	05:05
Control Test	0.080	05:05
Air Blank	0.000	05:06
Diagnostics Check	OK	05:06

Data Download Date/Time: 02/27/2019 09:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001075

INSTRUMENT REGISTERED TO: FHP TROOP E

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CAPOTE,DUVAN,

BREATH TEST OPERATOR AGENCY: FHP Troop E

ARRESTING OFFICER: CAPOTE,DUVAN

ARRESTING OFFICER AGENCY: FHP Troop E

LAST AGENCY INSPECTION: 01/02/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 712031

DRY GAS EXPIRY DATE: 02/21/2020

VIOLATION CODE: DUI

ARREST TIME: 22:17

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: BRISSETT

DATE/TIME OF TEST: 02/08/2019 23:24:33

Diagnostics Check	OK	23:18
Air Blank	0.000	23:18
Control Test	0.081	23:19
Air Blank	0.000	23:19
Subject Sample # 1	0.178	23:20
Breath Volume	1.949	23:20
Air Blank	0.000	23:20
Air Blank	0.000	23:22
Subject Sample # 2	0.170	23:23
Breath Volume	2.480	23:23
Air Blank	0.000	23:23
Control Test	0.080	23:23
Air Blank	0.000	23:24
Diagnostics Check	OK	23:24

Data Download Date/Time: 02/27/2019 09:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001075

INSTRUMENT REGISTERED TO: FHP TROOP E

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GARCIA,HENRY,E

BREATH TEST OPERATOR AGENCY: FHP Troop E

ARRESTING OFFICER: GARCIA,HENRY

ARRESTING OFFICER AGENCY: FHP Troop E

LAST AGENCY INSPECTION: 01/02/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 712031

DRY GAS EXPIRY DATE: 02/21/2020

VIOLATION CODE: DUI

ARREST TIME: 17:05

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: BESWICK

DATE/TIME OF TEST: 02/11/2019 20:25:30

Diagnostics Check	OK	20:18
Air Blank	0.000	20:19
Control Test	0.080	20:19
Air Blank	0.000	20:19
Subject Sample # 1	0.186	20:20
Breath Volume	1.128	20:20
Air Blank	0.000	20:20
Air Blank	0.000	20:22
Subject Sample # 2	0.193	20:24
Breath Volume	1.332	20:24
Air Blank	0.000	20:24
Control Test	0.080	20:24
Air Blank	0.000	20:25
Diagnostics Check	OK	20:25

Data Download Date/Time: 02/27/2019 09:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001075

INSTRUMENT REGISTERED TO: FHP TROOP E

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: REINOSO,PEDRO,

BREATH TEST OPERATOR AGENCY: FHP Troop E

ARRESTING OFFICER: GIZ,JOAQUIN

ARRESTING OFFICER AGENCY: FHP Troop E

LAST AGENCY INSPECTION: 01/02/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 712031

DRY GAS EXPIRY DATE: 02/21/2020

VIOLATION CODE: DUI

ARREST TIME: 03:35

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: RODRIGUEZ GARCIA

DATE/TIME OF TEST: 02/17/2019 03:59:11

Diagnostics Check	OK	03:56	
Air Blank	0.000	03:56	
Control Test	0.080	03:57	
Air Blank	0.000	03:57	
Subject Sample	# 1	03:57	Subject Test Refused
Breath Volume	0.000	03:57	
Air Blank	0.000	03:58	
Control Test	0.080	03:58	
Air Blank	0.000	03:59	
Diagnostics Check	OK	03:59	

Data Download Date/Time: 02/27/2019 09:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001075

INSTRUMENT REGISTERED TO: FHP TROOP E

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SIERRA,ALIEK,O

BREATH TEST OPERATOR AGENCY: FHP Troop E

ARRESTING OFFICER: SIERRA,ALIEK

ARRESTING OFFICER AGENCY: FHP Troop E

LAST AGENCY INSPECTION: 01/02/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 712031

DRY GAS EXPIRY DATE: 02/21/2020

VIOLATION CODE: DUI

ARREST TIME: 21:52

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: GONZALEZ MONTANO

DATE/TIME OF TEST: 02/24/2019 22:55:34

Diagnostics Check	OK	22:52	
Air Blank	0.000	22:53	
Control Test	0.081	22:53	
Air Blank	0.000	22:53	
Subject Sample	# 1	22:54	Subject Test Refused
Breath Volume	0.000	22:54	
Air Blank	0.000	22:54	
Control Test	0.081	22:54	
Air Blank	0.000	22:55	
Diagnostics Check	OK	22:55	

Data Download Date/Time: 02/27/2019 09:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER,JOHN,T
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: HIPPLE,ALEX
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 02/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 18:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: JIMINEZ

DATE/TIME OF TEST: 02/27/2019 18:52:42

Diagnostics Check	OK	18:46
Air Blank	0.000	18:46
Control Test	0.078	18:46
Air Blank	0.000	18:47
Subject Sample # 1	0.267	18:48
Breath Volume	2.917	18:48
Air Blank	0.000	18:48
Air Blank	0.000	18:50
Subject Sample # 2	0.271	18:51
Breath Volume	2.257	18:51
Air Blank	0.000	18:51
Control Test	0.078	18:52
Air Blank	0.000	18:52
Diagnostics Check	OK	18:52

Data Download Date/Time: 03/31/2019 12:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CRUM,ZACKARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: MAAS

DATE/TIME OF TEST: 02/01/2019 23:14:35

Diagnostics Check	OK	23:11	
Air Blank	0.000	23:11	
Control Test	0.080	23:12	
Air Blank	0.000	23:12	
Subject Sample	# 1 0.189	23:13	
Breath Volume	3.890	23:13	
Air Blank		23:13	Purge Fail
Air Blank	0.000	23:14	

Data Download Date/Time: 02/06/2019 15:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CRUM,ZACKARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: MAAS

DATE/TIME OF TEST: 02/02/2019 00:03:40

Diagnostics Check	OK	00:00	
Air Blank	0.000	00:00	
Control Test	0.076	00:01	
Air Blank	0.000	00:01	
Subject Sample	# 1 0.164	00:02	
Breath Volume	3.246	00:02	
Air Blank		00:02	Purge Fail
Air Blank		00:03	Purge Fail

Data Download Date/Time: 02/06/2019 15:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CRUM,ZACKARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: MAAS

DATE/TIME OF TEST: 02/02/2019 00:08:58

Diagnostics Check	OK	00:07	
Air Blank	0.000	00:08	
Control Test	0.054	00:08	Control Outside Tolerance
Air Blank	0.000	00:08	

Data Download Date/Time: 02/06/2019 15:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CRUM,ZACKARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: MAAS

DATE/TIME OF TEST: 02/02/2019 00:12:02

Diagnostics Check	OK	00:10	
Air Blank	0.000	00:11	
Control Test	0.065	00:11	Control Outside Tolerance
Air Blank	0.000	00:12	

Data Download Date/Time: 02/06/2019 15:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUM,ZACKARY,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KARAYIANES,M
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 02/02/2019 00:37:55

Diagnostics Check	OK	00:36	
Air Blank	0.000	00:37	
Control Test	0.071	00:37	Control Outside Tolerance
Air Blank	0.000	00:37	

Data Download Date/Time: 02/06/2019 15:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUM,ZACKARY,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KARAYIANES,M
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 02/02/2019 01:12:56

Diagnostics Check	OK	01:06
Air Blank	0.000	01:07
Control Test	0.075	01:07
Air Blank	0.000	01:07
Subject Sample # 1	0.000	01:08
Breath Volume	3.335	01:08
Air Blank	0.000	01:09
Air Blank	0.000	01:11
Subject Sample # 2	0.000	01:11
Breath Volume	2.664	01:11
Air Blank	0.000	01:12
Control Test	0.076	01:12
Air Blank	0.000	01:12
Diagnostics Check	OK	01:12

Data Download Date/Time: 02/06/2019 15:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CRUM,ZACKARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: MAAS

DATE/TIME OF TEST: 02/02/2019 01:29:15

Diagnostics Check	OK	01:16	
Air Blank	0.000	01:17	
Control Test	0.076	01:17	
Air Blank	0.000	01:18	
Subject Sample # 1	0.131	01:18	
Breath Volume	1.238	01:18	
Air Blank	0.000	01:19	
Air Blank	0.000	01:21	
Subject Sample # 2		01:24	No Sample Provided
Breath Volume	0.000	01:24	
Air Blank	0.000	01:25	
Air Blank	0.000	01:27	
Subject Sample # 3		01:27	Subject Test Refused
Breath Volume	0.000	01:27	
Air Blank	0.000	01:27	
Control Test	0.076	01:27	
Air Blank	0.000	01:28	
Diagnostics Check	OK	01:29	

Data Download Date/Time: 02/06/2019 15:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISTI,JORDAN,A
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SISTI,JORDAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:58

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 02/02/2019 05:20:02

Diagnostics Check	OK	05:18	
Air Blank	0.000	05:19	
Control Test	0.072	05:19	Control Outside Tolerance
Air Blank	0.000	05:20	

Data Download Date/Time: 02/06/2019 15:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISTI,JORDAN,A
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SISTI,JORDAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:58

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 02/02/2019 05:32:06

Diagnostics Check	OK	05:29	
Air Blank	0.000	05:29	
Control Test	0.076	05:30	
Air Blank	0.000	05:30	
Subject Sample	# 1	05:30	Subject Test Refused
Breath Volume	0.000	05:30	
Air Blank	0.000	05:31	
Control Test	0.078	05:31	
Air Blank	0.000	05:32	
Diagnostics Check	OK	05:32	

Data Download Date/Time: 02/06/2019 15:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISTI,JORDAN,A
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BUTTERFIELD,KELLEY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: ZEWEN

DATE/TIME OF TEST: 02/05/2019 02:24:10

Diagnostics Check	OK	02:20	
Air Blank	0.000	02:21	
Control Test	0.081	02:21	
Air Blank	0.000	02:22	
Subject Sample	# 1	02:22	Subject Test Refused
Breath Volume	0.000	02:22	
Air Blank	0.000	02:23	
Control Test	0.080	02:23	
Air Blank	0.000	02:24	
Diagnostics Check	OK	02:24	

Data Download Date/Time: 02/06/2019 15:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLINE,NICOLE,S
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: RUIZ,CESAR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: FLAMURI

DATE/TIME OF TEST: 02/06/2019 01:46:31

Diagnostics Check	OK	01:35	
Air Blank	0.000	01:35	
Control Test	0.080	01:35	
Air Blank	0.000	01:36	
Subject Sample	# 1	01:39	No Sample Provided
Breath Volume	0.000	01:39	
Air Blank	0.000	01:40	
Air Blank	0.000	01:42	
Subject Sample	# 2	01:45	No Sample Provided
Breath Volume	0.000	01:45	
Air Blank	0.000	01:45	
Control Test	0.081	01:45	
Air Blank	0.000	01:46	
Diagnostics Check	OK	01:46	

Data Download Date/Time: 02/06/2019 15:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRONZ,BRIAN,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,KRISTEN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:32

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 02/07/2019 20:39:38

Diagnostics Check	OK	20:37	
Air Blank	0.000	20:37	
Control Test	0.081	20:38	
Air Blank	0.000	20:38	
Subject Sample	# 1	20:39	Sequence Aborted
Breath Volume	0.000	20:39	
Air Blank	0.000	20:39	

Data Download Date/Time: 03/07/2019 09:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRONZ,BRIAN,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,KRISTEN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:32

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 02/07/2019 20:58:43

Diagnostics Check	OK	20:52
Air Blank	0.000	20:53
Control Test	0.081	20:53
Air Blank	0.000	20:54
Subject Sample # 1	0.000	20:54
Breath Volume	2.843	20:54
Air Blank	0.000	20:54
Air Blank	0.000	20:56
Subject Sample # 2	0.000	20:57
Breath Volume	3.785	20:57
Air Blank	0.000	20:57
Control Test	0.081	20:58
Air Blank	0.000	20:58
Diagnostics Check	OK	20:58

Data Download Date/Time: 03/07/2019 09:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAIA,MICHAEL,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WOOLVERTON,JUSTIN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 02/08/2019 04:33:24

Diagnostics Check	OK	04:24	
Air Blank	0.000	04:25	
Control Test	0.080	04:25	
Air Blank	0.000	04:25	
Subject Sample	# 1	04:29	No Sample Provided
Breath Volume	0.000	04:29	
Air Blank	0.000	04:29	
Air Blank	0.000	04:31	
Subject Sample	# 2	04:32	Subject Test Refused
Breath Volume	0.000	04:32	
Air Blank	0.000	04:32	
Control Test	0.080	04:32	
Air Blank	0.000	04:33	
Diagnostics Check	OK	04:33	

Data Download Date/Time: 03/07/2019 09:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE, RONALD, W
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCKENZIE, RONALD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: WEBER

DATE/TIME OF TEST: 02/08/2019 17:17:38

Diagnostics Check	OK	17:11
Air Blank	0.000	17:11
Control Test	0.081	17:12
Air Blank	0.000	17:12
Subject Sample # 1	0.111	17:13
Breath Volume	1.281	17:13
Air Blank	0.000	17:13
Air Blank	0.000	17:15
Subject Sample # 2	0.111	17:16
Breath Volume	1.390	17:16
Air Blank	0.000	17:16
Control Test	0.079	17:17
Air Blank	0.000	17:17
Diagnostics Check	OK	17:17

Data Download Date/Time: 03/07/2019 09:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAIA,MICHAEL,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HOBART,CLAYTON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: BROWNLEE

DATE/TIME OF TEST: 02/09/2019 04:42:47

Diagnostics Check	OK	04:31	
Air Blank	0.000	04:31	
Control Test	0.081	04:32	
Air Blank	0.000	04:32	
Subject Sample	# 1 0.033	04:35	Volume Not Met
Breath Volume	0.082	04:35	
Air Blank	0.000	04:36	
Air Blank	0.000	04:38	
Subject Sample	# 2	04:41	Subject Test Refused
Breath Volume	0.019	04:41	
Air Blank	0.000	04:41	
Control Test	0.078	04:42	
Air Blank	0.000	04:42	
Diagnostics Check	OK	04:42	

Data Download Date/Time: 03/07/2019 09:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUM,ZACKARY,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KARAYIANES,M
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:09

SUBJECT'S LAST NAME: SAYTHANOM

DATE/TIME OF TEST: 02/09/2019 20:29:43

Diagnostics Check	OK	20:23
Air Blank	0.000	20:24
Control Test	0.081	20:24
Air Blank	0.000	20:24
Subject Sample # 1	0.185	20:25
Breath Volume	2.210	20:25
Air Blank	0.000	20:25
Air Blank	0.000	20:27
Subject Sample # 2	0.184	20:28
Breath Volume	2.296	20:28
Air Blank	0.000	20:28
Control Test	0.080	20:29
Air Blank	0.000	20:29
Diagnostics Check	OK	20:29

Data Download Date/Time: 03/07/2019 09:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LAMOUR,ZACHARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:18

SUBJECT'S LAST NAME: RAGSDALE

DATE/TIME OF TEST: 02/09/2019 21:02:34

Diagnostics Check	OK	20:56
Air Blank	0.000	20:56
Control Test	0.080	20:57
Air Blank	0.000	20:57
Subject Sample # 1	0.272	20:57
Breath Volume	2.886	20:57
Air Blank	0.000	20:58
Air Blank	0.000	21:00
Subject Sample # 2	0.278	21:00
Breath Volume	3.128	21:00
Air Blank	0.000	21:01
Control Test	0.079	21:01
Air Blank	0.000	21:02
Diagnostics Check	OK	21:02

Data Download Date/Time: 03/07/2019 09:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CRUM,ZACKARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:28

SUBJECT'S LAST NAME: LOVETT

DATE/TIME OF TEST: 02/09/2019 22:03:00

Diagnostics Check	OK	21:56
Air Blank	0.000	21:56
Control Test	0.081	21:57
Air Blank	0.000	21:57
Subject Sample # 1	0.074	21:58
Breath Volume	1.601	21:58
Air Blank	0.000	21:59
Air Blank	0.000	22:01
Subject Sample # 2	0.067	22:01
Breath Volume	1.433	22:01
Air Blank	0.000	22:02
Control Test	0.079	22:02
Air Blank	0.000	22:02
Diagnostics Check	OK	22:03

Data Download Date/Time: 03/07/2019 09:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUM,ZACKARY,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KARAYIANES,M
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: RANDOLPH

DATE/TIME OF TEST: 02/10/2019 01:14:07

Diagnostics Check	OK	01:08
Air Blank	0.000	01:08
Control Test	0.081	01:08
Air Blank	0.000	01:09
Subject Sample # 1	0.000	01:09
Breath Volume	2.000	01:09
Air Blank	0.000	01:10
Air Blank	0.000	01:12
Subject Sample # 2	0.000	01:12
Breath Volume	3.039	01:12
Air Blank	0.000	01:13
Control Test	0.080	01:13
Air Blank	0.000	01:14
Diagnostics Check	OK	01:14

Data Download Date/Time: 03/07/2019 09:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LAMOUR,ZACHARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: SANDOVAL

DATE/TIME OF TEST: 02/10/2019 02:37:39

Diagnostics Check	OK	02:29
Air Blank	0.000	02:30
Control Test	0.081	02:30
Air Blank	0.000	02:30
Subject Sample # 1	0.225	02:32
Breath Volume	1.664	02:32
Air Blank	0.000	02:33
Air Blank	0.000	02:35
Subject Sample # 2	0.218	02:36
Breath Volume	1.921	02:36
Air Blank	0.000	02:36
Control Test	0.079	02:37
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

Data Download Date/Time: 03/07/2019 09:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WOOLVERTON,JUSTIN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: URDAL

DATE/TIME OF TEST: 02/10/2019 03:24:48

Diagnostics Check	OK	03:18
Air Blank	0.000	03:19
Control Test	0.079	03:19
Air Blank	0.000	03:19
Subject Sample # 1	0.172	03:20
Breath Volume	1.929	03:20
Air Blank	0.000	03:20
Air Blank	0.000	03:22
Subject Sample # 2	0.182	03:23
Breath Volume	1.984	03:23
Air Blank	0.000	03:23
Control Test	0.079	03:24
Air Blank	0.000	03:24
Diagnostics Check	OK	03:24

Data Download Date/Time: 03/07/2019 09:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCE JR,SCOTT,A
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTTA,VLADAMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:41

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 02/13/2019 16:04:51

Diagnostics Check	OK	16:01	
Air Blank	0.000	16:02	
Control Test	0.081	16:02	
Air Blank	0.000	16:02	
Subject Sample	# 1	16:03	Subject Test Refused
Breath Volume	0.000	16:03	
Air Blank	0.000	16:03	
Control Test	0.081	16:04	
Air Blank	0.000	16:04	
Diagnostics Check	OK	16:04	

Data Download Date/Time: 03/07/2019 09:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCE JR,SCOTT,A
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTTA,VLADAMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:12

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 02/13/2019 16:15:32

Diagnostics Check	OK	16:14	
Air Blank		16:15	Sequence Aborted
Air Blank	0.000	16:15	

Data Download Date/Time: 03/07/2019 09:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCE JR,SCOTT,A
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTTA,VLADAMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:41

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 02/13/2019 16:23:12

Diagnostics Check	OK	16:17	
Air Blank	0.000	16:18	
Control Test	0.081	16:18	
Air Blank	0.000	16:18	
Subject Sample	# 1	16:21	Subject Test Refused
Breath Volume	0.574	16:21	
Air Blank	0.000	16:22	
Control Test	0.081	16:22	
Air Blank	0.000	16:23	
Diagnostics Check	OK	16:23	

Data Download Date/Time: 03/07/2019 09:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLINE,NICOLE,S
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WOOLVERTON,JUSTIN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: SCOTTI

DATE/TIME OF TEST: 02/14/2019 03:27:07

Diagnostics Check	OK	03:21
Air Blank	0.000	03:21
Control Test	0.080	03:21
Air Blank	0.000	03:22
Subject Sample # 1	0.127	03:22
Breath Volume	3.480	03:22
Air Blank	0.000	03:23
Air Blank	0.000	03:25
Subject Sample # 2	0.129	03:25
Breath Volume	3.777	03:25
Air Blank	0.000	03:26
Control Test	0.082	03:26
Air Blank	0.000	03:27
Diagnostics Check	OK	03:27

Data Download Date/Time: 03/07/2019 09:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUM,ZACKARY,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KARAYIANES,M
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 02/15/2019 00:50:43

Diagnostics Check	OK	00:36	
Air Blank	0.000	00:37	
Control Test	0.081	00:37	
Air Blank	0.000	00:38	
Subject Sample # 1	0.000	00:40	
Breath Volume	1.757	00:40	
Air Blank	0.000	00:40	
Air Blank	0.000	00:42	
Subject Sample # 2	0.000	00:45	Volume Not Met
Breath Volume	0.792	00:45	
Air Blank	0.000	00:46	
Air Blank	0.000	00:48	
Subject Sample # 3	0.000	00:48	
Breath Volume	1.109	00:48	
Air Blank	0.000	00:49	
Control Test	0.080	00:49	
Air Blank	0.000	00:50	
Diagnostics Check	Fail	00:50	Voltage/Current Test
Air Blank	0.000	00:50	

Data Download Date/Time: 03/07/2019 09:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUM,ZACKARY,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KARAYIANES,M
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 02/15/2019 01:01:01

Diagnostics Check	OK	00:52	
Air Blank	0.000	00:53	
Control Test	0.081	00:53	
Air Blank	0.000	00:54	
Subject Sample # 1	0.000	00:55	
Breath Volume	1.515	00:55	
Air Blank	0.000	00:55	
Air Blank	0.000	00:57	
Subject Sample # 2	0.000	00:59	
Breath Volume	1.566	00:59	
Air Blank	0.000	00:59	
Control Test	0.081	00:59	
Air Blank	0.000	01:00	
Diagnostics Check	Fail	01:00	Voltage/Current Test
Air Blank	0.000	01:01	

Data Download Date/Time: 03/07/2019 09:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LESPERANCE,DANIEL,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LAMOUR,ZACHARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 02/18/2019 00:22:50

Diagnostics Check	OK	00:13
Air Blank	0.000	00:14
Control Test	0.081	00:14
Air Blank	0.000	00:14
Subject Sample # 1	0.100	00:15
Breath Volume	1.710	00:15
Air Blank	0.000	00:16
Air Blank	0.000	00:17
Subject Sample # 2	0.123	00:18
Breath Volume	2.105	00:18
Air Blank	0.000	00:19
Air Blank	0.000	00:20
Subject Sample # 3	0.123	00:21
Breath Volume	2.890	00:21
Air Blank	0.000	00:21
Control Test	0.079	00:22
Air Blank	0.000	00:22
Diagnostics Check	OK	00:22

Data Download Date/Time: 03/07/2019 09:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUM,ZACKARY,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LAMOUR,Z
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: LEACH

DATE/TIME OF TEST: 02/18/2019 03:35:11

Diagnostics Check	OK	03:31	
Air Blank	0.000	03:32	
Control Test	0.079	03:32	
Air Blank	0.000	03:33	
Subject Sample	# 1	03:33	Subject Test Refused
Breath Volume	0.000	03:33	
Air Blank	0.000	03:34	
Control Test	0.079	03:34	
Air Blank	0.000	03:35	
Diagnostics Check	OK	03:35	

Data Download Date/Time: 03/07/2019 09:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINES,PAUL,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HINES,PAUL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/06/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 02/18/2019 23:29:33

Diagnostics Check	OK	23:21	
Air Blank	0.000	23:21	
Control Test	0.080	23:21	
Air Blank	0.000	23:22	
Subject Sample	# 1	23:25	No Sample Provided
Breath Volume	0.000	23:25	
Air Blank	0.000	23:25	
Air Blank	0.000	23:27	
Subject Sample	# 2	23:28	Subject Test Refused
Breath Volume	0.000	23:28	
Air Blank	0.000	23:28	
Control Test	0.081	23:28	
Air Blank	0.000	23:29	
Diagnostics Check	OK	23:29	

Data Download Date/Time: 03/07/2019 09:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LAMOUR,ZACHARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: BAIZAN

DATE/TIME OF TEST: 02/21/2019 03:05:55

Diagnostics Check	OK	02:57	
Air Blank	0.000	02:57	
Control Test	0.079	02:57	
Air Blank	0.000	02:58	
Subject Sample	# 1	03:01	No Sample Provided
Breath Volume	0.000	03:01	
Air Blank	0.000	03:02	
Air Blank	0.000	03:04	
Subject Sample	# 2	03:04	Subject Test Refused
Breath Volume	0.000	03:04	
Air Blank	0.000	03:04	
Control Test	0.078	03:05	
Air Blank	0.000	03:05	
Diagnostics Check	OK	03:05	

Data Download Date/Time: 03/07/2019 09:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LESPERANCE,DANIEL,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LESPERANCE,DANIEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: CHERANICHIT

DATE/TIME OF TEST: 02/22/2019 01:12:39

Diagnostics Check	OK	01:06
Air Blank	0.000	01:07
Control Test	0.080	01:07
Air Blank	0.000	01:07
Subject Sample # 1	0.162	01:08
Breath Volume	1.261	01:08
Air Blank	0.000	01:08
Air Blank	0.000	01:10
Subject Sample # 2	0.166	01:11
Breath Volume	1.253	01:11
Air Blank	0.000	01:11
Control Test	0.079	01:12
Air Blank	0.000	01:12
Diagnostics Check	OK	01:12

Data Download Date/Time: 03/07/2019 09:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LESPERANCE,DANIEL,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CRUM,ZACHARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: AL QAISI

DATE/TIME OF TEST: 02/23/2019 00:40:05

Diagnostics Check	OK	00:33
Air Blank	0.000	00:33
Control Test	0.081	00:34
Air Blank	0.000	00:34
Subject Sample # 1	0.158	00:35
Breath Volume	1.343	00:35
Air Blank	0.000	00:36
Air Blank	0.000	00:38
Subject Sample # 2	0.161	00:38
Breath Volume	1.285	00:38
Air Blank	0.000	00:39
Control Test	0.080	00:39
Air Blank	0.000	00:39
Diagnostics Check	OK	00:40

Data Download Date/Time: 03/07/2019 09:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LESPERANCE,DANIEL,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KARAYIANES,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 02/23/2019 02:03:09

Diagnostics Check	OK	01:56
Air Blank	0.000	01:56
Control Test	0.081	01:57
Air Blank	0.000	01:57
Subject Sample # 1	0.245	01:58
Breath Volume	1.183	01:58
Air Blank	0.000	01:58
Air Blank	0.000	02:00
Subject Sample # 2	0.245	02:01
Breath Volume	1.382	02:01
Air Blank	0.000	02:02
Control Test	0.080	02:02
Air Blank	0.000	02:03
Diagnostics Check	OK	02:03

Data Download Date/Time: 03/07/2019 09:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINES,PAUL,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HINES,PAUL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: WHITLOCK

DATE/TIME OF TEST: 02/25/2019 02:29:58

Diagnostics Check	OK	02:21	
Air Blank	0.000	02:21	
Control Test	0.080	02:21	
Air Blank	0.000	02:22	
Subject Sample	# 1	02:25	No Sample Provided
Breath Volume	0.000	02:25	
Air Blank	0.000	02:25	
Air Blank	0.000	02:27	
Subject Sample	# 2	02:28	Subject Test Refused
Breath Volume	0.000	02:28	
Air Blank	0.000	02:29	
Control Test	0.081	02:29	
Air Blank	0.000	02:29	
Diagnostics Check	OK	02:29	

Data Download Date/Time: 03/07/2019 09:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLINE,NICOLE,S
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HINES,PAUL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: PATTON

DATE/TIME OF TEST: 02/25/2019 23:18:21

Diagnostics Check	OK	23:06	
Air Blank	0.000	23:07	
Control Test	0.081	23:07	
Air Blank	0.000	23:08	
Subject Sample	# 1	23:11	No Sample Provided
Breath Volume	0.000	23:11	
Air Blank	0.000	23:11	
Air Blank	0.000	23:13	
Subject Sample	# 2	23:16	No Sample Provided
Breath Volume	0.000	23:16	
Air Blank	0.000	23:17	
Control Test	0.080	23:17	
Air Blank	0.000	23:18	
Diagnostics Check	OK	23:18	

Data Download Date/Time: 03/07/2019 09:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUM,ZACKARY,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LAMOUR,Z
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:33

SUBJECT'S LAST NAME: WIGGINS

DATE/TIME OF TEST: 02/28/2019 22:04:50

Diagnostics Check	OK	21:57
Air Blank	0.000	21:57
Control Test	0.077	21:58
Air Blank	0.000	21:58
Subject Sample # 1	0.298	22:00
Breath Volume	3.203	22:00
Air Blank	0.000	22:00
Air Blank	0.000	22:02
Subject Sample # 2	0.283	22:03
Breath Volume	2.976	22:03
Air Blank	0.000	22:03
Control Test	0.077	22:04
Air Blank	0.000	22:04
Diagnostics Check	OK	22:04

Data Download Date/Time: 03/07/2019 09:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: SILVERSTEIN,CHRISTINA

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 01/23/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17817080A3

DRY GAS EXPIRY DATE: 08/05/2019

VIOLATION CODE: DUI

ARREST TIME: 23:28

GENDER: Female

OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 02/01/2019 00:29:22

Diagnostics Check	OK	00:17	
Air Blank	0.000	00:17	
Control Test	0.080	00:18	
Air Blank	0.000	00:18	
Subject Sample	# 1 0.127	00:21	Volume Not Met
Breath Volume	1.011	00:21	
Air Blank	0.000	00:22	
Air Blank	0.000	00:24	
Subject Sample	# 2 0.144	00:24	
Breath Volume	1.980	00:24	
Air Blank	0.000	00:25	
Air Blank	0.000	00:27	
Subject Sample	# 3 0.145	00:27	
Breath Volume	2.433	00:27	
Air Blank	0.000	00:28	
Control Test	0.079	00:28	
Air Blank	0.000	00:29	
Diagnostics Check	OK	00:29	

Data Download Date/Time: 02/06/2019 13:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: SILVERSTEIN,CHRISTINA

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 01/23/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17817080A3

DRY GAS EXPIRY DATE: 08/05/2019

VIOLATION CODE: DUI

ARREST TIME: 23:17

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: ECKENSBERGER

DATE/TIME OF TEST: 02/02/2019 00:02:12

Diagnostics Check	OK	23:52
Air Blank	0.000	23:53
Control Test	0.079	23:53
Air Blank	0.000	23:54
Subject Sample # 2	0.150	23:57
Breath Volume	1.613	23:57
Air Blank	0.000	23:57
Air Blank	0.000	23:59
Subject Sample # 1	0.158	00:00
Breath Volume	1.335	00:00
Air Blank	0.000	00:01
Control Test	0.079	00:01
Air Blank	0.000	00:02
Diagnostics Check	OK	00:02

Data Download Date/Time: 02/06/2019 13:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HASTINGS,KENNETH,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: HASTINGS,KENNETH

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 02/06/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17817080A3

DRY GAS EXPIRY DATE: 08/05/2019

VIOLATION CODE: DUI

ARREST TIME: 23:17

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: BOOKWALTER

DATE/TIME OF TEST: 02/06/2019 23:59:23

Diagnostics Check	OK	23:56	
Air Blank	0.000	23:56	
Control Test	0.080	23:56	
Air Blank	0.000	23:57	
Subject Sample	# 1	23:58	Subject Test Refused
Breath Volume	0.000	23:58	
Air Blank	0.000	23:58	
Control Test	0.079	23:58	
Air Blank	0.000	23:59	
Diagnostics Check	OK	23:59	

Data Download Date/Time: 03/12/2019 08:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 02/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 02/08/2019 04:23:34

Diagnostics Check	OK	04:17
Air Blank	0.000	04:17
Control Test	0.079	04:18
Air Blank	0.000	04:18
Subject Sample # 1	0.150	04:19
Breath Volume	2.652	04:19
Air Blank	0.000	04:19
Air Blank	0.000	04:21
Subject Sample # 2	0.147	04:22
Breath Volume	2.000	04:22
Air Blank	0.000	04:22
Control Test	0.080	04:22
Air Blank	0.000	04:23
Diagnostics Check	OK	04:23

Data Download Date/Time: 03/12/2019 08:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 02/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: STABINS

DATE/TIME OF TEST: 02/09/2019 02:16:24

Diagnostics Check	OK	02:08
Air Blank	0.000	02:08
Control Test	0.080	02:09
Air Blank	0.000	02:09
Subject Sample # 1	0.186	02:11
Breath Volume	2.343	02:11
Air Blank	0.000	02:11
Air Blank	0.000	02:13
Subject Sample # 2	0.180	02:14
Breath Volume	1.777	02:14
Air Blank	0.000	02:15
Control Test	0.079	02:15
Air Blank	0.000	02:16
Diagnostics Check	OK	02:16

Data Download Date/Time: 03/12/2019 08:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HASTINGS,KENNETH,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: HASTINGS,KENNETH

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 02/06/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17817080A3

DRY GAS EXPIRY DATE: 08/05/2019

VIOLATION CODE: DUI

ARREST TIME: 03:46

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: MYRICK

DATE/TIME OF TEST: 02/09/2019 04:32:43

Diagnostics Check	OK	04:24	
Air Blank	0.000	04:24	
Control Test	0.079	04:24	
Air Blank	0.000	04:25	
Subject Sample	# 1	04:28	No Sample Provided
Breath Volume	0.000	04:28	
Air Blank	0.000	04:28	
Air Blank	0.000	04:30	
Subject Sample	# 2	04:31	Subject Test Refused
Breath Volume	0.000	04:31	
Air Blank	0.000	04:31	
Control Test	0.079	04:32	
Air Blank	0.000	04:32	
Diagnostics Check	OK	04:32	

Data Download Date/Time: 03/12/2019 08:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 02/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: RAY

DATE/TIME OF TEST: 02/10/2019 02:41:42

Diagnostics Check	OK	02:36	
Air Blank	0.000	02:36	
Control Test	0.079	02:37	
Air Blank	0.000	02:37	
Subject Sample	# 1	02:40	Subject Test Refused
Breath Volume	0.000	02:40	
Air Blank	0.000	02:40	
Control Test	0.080	02:41	
Air Blank	0.000	02:41	
Diagnostics Check	OK	02:41	

Data Download Date/Time: 03/12/2019 08:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 02/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: BECK

DATE/TIME OF TEST: 02/14/2019 04:01:49

Diagnostics Check	OK	03:47	
Air Blank	0.000	03:48	
Control Test	0.080	03:48	
Air Blank	0.000	03:48	
Subject Sample	# 1 0.078	03:51	
Breath Volume	1.152	03:51	
Air Blank	0.000	03:51	
Air Blank	0.000	03:53	
Subject Sample	# 2 0.076	03:57	Slope Not Level
Breath Volume	1.164	03:57	
Air Blank	0.000	03:57	
Air Blank	0.000	03:59	
Subject Sample	# 3 0.083	04:00	
Breath Volume	1.664	04:00	
Air Blank	0.000	04:00	
Control Test	0.079	04:01	
Air Blank	0.000	04:01	
Diagnostics Check	OK	04:01	

Data Download Date/Time: 03/12/2019 08:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KATZ,JUSTIN,T
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SIMS,ALEXANDRIA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 02/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:14

SUBJECT'S LAST NAME: FENDER

DATE/TIME OF TEST: 02/14/2019 17:45:55

Diagnostics Check	OK	17:34
Air Blank	0.000	17:35
Control Test	0.080	17:35
Air Blank	0.000	17:35
Subject Sample # 1	0.234	17:36
Breath Volume	1.683	17:36
Air Blank	0.000	17:37
Air Blank	0.000	17:39
Subject Sample # 2	0.213	17:40
Breath Volume	1.738	17:40
Air Blank	0.000	17:41
Air Blank	0.000	17:43
Subject Sample # 3	0.221	17:44
Breath Volume	1.187	17:44
Air Blank	0.000	17:44
Control Test	0.079	17:45
Air Blank	0.000	17:45
Diagnostics Check	OK	17:45

Data Download Date/Time: 03/12/2019 08:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 02/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: BEASLEY-KANASH

DATE/TIME OF TEST: 02/15/2019 03:25:51

Diagnostics Check	OK	03:18
Air Blank	0.000	03:19
Control Test	0.080	03:19
Air Blank	0.000	03:19
Subject Sample # 1	0.051	03:20
Breath Volume	2.234	03:20
Air Blank	0.000	03:21
Air Blank	0.000	03:23
Subject Sample # 2	0.052	03:24
Breath Volume	2.332	03:24
Air Blank	0.000	03:24
Control Test	0.079	03:25
Air Blank	0.000	03:25
Diagnostics Check	OK	03:25

Data Download Date/Time: 03/12/2019 08:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HASTINGS,KENNETH,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: HASTINGS,KENNETH

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 02/06/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17817080A3

DRY GAS EXPIRY DATE: 08/05/2019

VIOLATION CODE: DUI

ARREST TIME: 21:55

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: NAVA CRUZ

DATE/TIME OF TEST: 02/15/2019 22:44:29

Diagnostics Check	OK	22:38
Air Blank	0.000	22:38
Control Test	0.079	22:39
Air Blank	0.000	22:39
Subject Sample # 1	0.149	22:40
Breath Volume	3.781	22:40
Air Blank	0.000	22:40
Air Blank	0.000	22:42
Subject Sample # 2	0.148	22:42
Breath Volume	4.101	22:42
Air Blank	0.000	22:43
Control Test	0.078	22:43
Air Blank	0.000	22:44
Diagnostics Check	OK	22:44

Data Download Date/Time: 03/12/2019 08:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 02/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: DELEEUEW

DATE/TIME OF TEST: 02/16/2019 02:05:49

Diagnostics Check	OK	01:56
Air Blank	0.000	01:56
Control Test	0.079	01:56
Air Blank	0.000	01:57
Subject Sample # 1	0.222	01:58
Breath Volume	3.343	01:58
Air Blank	0.000	01:59
Air Blank	0.000	02:00
Subject Sample # 2	0.199	02:01
Breath Volume	1.101	02:01
Air Blank	0.000	02:01
Air Blank	0.000	02:03
Subject Sample # 3	0.188	02:04
Breath Volume	1.355	02:04
Air Blank	0.000	02:04
Control Test	0.079	02:05
Air Blank	0.000	02:05
Diagnostics Check	OK	02:05

Data Download Date/Time: 03/12/2019 08:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 02/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 02/17/2019 04:28:25

Diagnostics Check	OK	04:21
Air Blank	0.000	04:22
Control Test	0.080	04:22
Air Blank	0.000	04:22
Subject Sample # 1	0.178	04:23
Breath Volume	3.554	04:23
Air Blank	0.000	04:24
Air Blank	0.000	04:26
Subject Sample # 2	0.182	04:26
Breath Volume	3.457	04:26
Air Blank	0.000	04:27
Control Test	0.078	04:27
Air Blank	0.000	04:28
Diagnostics Check	OK	04:28

Data Download Date/Time: 03/12/2019 08:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HASTINGS,KENNETH,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: HASTINGS,KENNETH

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 02/06/2019

Date or Time Changed: Yes

DRY GAS STANDARD LOT #: 17817080A3

DRY GAS EXPIRY DATE: 08/05/2019

VIOLATION CODE: DUI

ARREST TIME: 20:12

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:36

SUBJECT'S LAST NAME: WILLS

DATE/TIME OF TEST: 02/20/2019 21:03:36

Diagnostics Check	OK	20:57
Air Blank	0.000	20:57
Control Test	0.080	20:58
Air Blank	0.000	20:58
Subject Sample # 1	0.000	20:59
Breath Volume	3.105	20:59
Air Blank	0.000	20:59
Air Blank	0.000	21:01
Subject Sample # 2	0.000	21:02
Breath Volume	2.269	21:02
Air Blank	0.000	21:02
Control Test	0.079	21:02
Air Blank	0.000	21:03
Diagnostics Check	OK	21:03

Data Download Date/Time: 03/12/2019 08:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 02/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: ROTHMANN

DATE/TIME OF TEST: 02/21/2019 01:42:23

Diagnostics Check	OK	01:34
Air Blank	0.000	01:34
Control Test	0.079	01:35
Air Blank	0.000	01:35
Subject Sample # 1	0.111	01:37
Breath Volume	2.203	01:37
Air Blank	0.000	01:38
Air Blank	0.000	01:40
Subject Sample # 2	0.099	01:40
Breath Volume	2.199	01:40
Air Blank	0.000	01:41
Control Test	0.079	01:41
Air Blank	0.000	01:42
Diagnostics Check	OK	01:42

Data Download Date/Time: 03/12/2019 08:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 02/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 02/22/2019 22:41:00

Diagnostics Check	OK	22:29
Air Blank	0.000	22:30
Control Test	0.080	22:30
Air Blank	0.000	22:31
Subject Sample # 1	0.107	22:32
Breath Volume	1.406	22:32
Air Blank	0.000	22:32
Air Blank	0.000	22:34
Subject Sample # 2	0.133	22:35
Breath Volume	2.136	22:35
Air Blank	0.000	22:36
Air Blank	0.000	22:38
Subject Sample # 3	0.127	22:39
Breath Volume	2.437	22:39
Air Blank	0.000	22:40
Control Test	0.079	22:40
Air Blank	0.000	22:40
Diagnostics Check	OK	22:41

Data Download Date/Time: 03/12/2019 08:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HASTINGS,KENNETH,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: HASTINGS,KENNETH

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 02/06/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17817080A3

DRY GAS EXPIRY DATE: 08/05/2019

VIOLATION CODE: DUI

ARREST TIME: 23:51

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: COVELL

DATE/TIME OF TEST: 02/23/2019 00:35:04

Diagnostics Check	OK	00:32	
Air Blank	0.000	00:32	
Control Test	0.080	00:32	
Air Blank	0.000	00:33	
Subject Sample	# 1	00:33	Subject Test Refused
Breath Volume	0.000	00:33	
Air Blank	0.000	00:34	
Control Test	0.080	00:34	
Air Blank	0.000	00:34	
Diagnostics Check	OK	00:35	

Data Download Date/Time: 03/12/2019 08:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 02/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: LANE

DATE/TIME OF TEST: 02/23/2019 04:35:22

Diagnostics Check	OK	04:28
Air Blank	0.000	04:29
Control Test	0.079	04:29
Air Blank	0.000	04:29
Subject Sample # 1	0.165	04:30
Breath Volume	3.273	04:30
Air Blank	0.000	04:31
Air Blank	0.000	04:32
Subject Sample # 2	0.157	04:33
Breath Volume	2.312	04:33
Air Blank	0.000	04:34
Control Test	0.079	04:34
Air Blank	0.000	04:35
Diagnostics Check	OK	04:35

Data Download Date/Time: 03/12/2019 08:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HASTINGS,KENNETH,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: HASTINGS,KENNETH

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 02/06/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17817080A3

DRY GAS EXPIRY DATE: 08/05/2019

VIOLATION CODE: DUI

ARREST TIME: 03:49

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: CARR

DATE/TIME OF TEST: 02/23/2019 05:18:11

Diagnostics Check	OK	05:06	
Air Blank	0.000	05:06	
Control Test	0.079	05:07	
Air Blank	0.000	05:07	
Subject Sample	# 1 0.203	05:10	Volume Not Met
Breath Volume	0.164	05:10	
Air Blank	0.000	05:11	
Air Blank	0.000	05:13	
Subject Sample	# 2 0.202	05:13	
Breath Volume	2.562	05:13	
Air Blank	0.000	05:14	
Air Blank	0.000	05:16	
Subject Sample	# 3 0.198	05:16	
Breath Volume	1.992	05:16	
Air Blank	0.000	05:17	
Control Test	0.078	05:17	
Air Blank	0.000	05:18	
Diagnostics Check	OK	05:18	

Data Download Date/Time: 03/12/2019 08:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: SILVERSTEIN,CHRISTINA

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 02/06/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17817080A3

DRY GAS EXPIRY DATE: 08/05/2019

VIOLATION CODE: DUI

ARREST TIME: 00:23

GENDER: Female

OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: REED

DATE/TIME OF TEST: 02/24/2019 01:09:37

Diagnostics Check	OK	01:03
Air Blank	0.000	01:03
Control Test	0.080	01:04
Air Blank	0.000	01:04
Subject Sample # 1	0.144	01:05
Breath Volume	2.656	01:05
Air Blank	0.000	01:05
Air Blank	0.000	01:07
Subject Sample # 2	0.147	01:08
Breath Volume	2.687	01:08
Air Blank	0.000	01:08
Control Test	0.078	01:09
Air Blank	0.000	01:09
Diagnostics Check	OK	01:09

Data Download Date/Time: 03/12/2019 08:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HASTINGS,KENNETH,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: HASTINGS,KENNETH

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 02/06/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17817080A3

DRY GAS EXPIRY DATE: 08/05/2019

VIOLATION CODE: DUI

ARREST TIME: 01:48

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: BASULTO

DATE/TIME OF TEST: 02/24/2019 02:45:35

Diagnostics Check	OK	02:38
Air Blank	0.000	02:38
Control Test	0.079	02:39
Air Blank	0.000	02:39
Subject Sample # 1	0.169	02:40
Breath Volume	2.167	02:40
Air Blank	0.000	02:41
Air Blank	0.000	02:42
Subject Sample # 2	0.165	02:44
Breath Volume	2.433	02:44
Air Blank	0.000	02:44
Control Test	0.079	02:44
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 03/12/2019 08:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICE,JOHN,B
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: RICE,JOHN
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 02/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 13:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:16

SUBJECT'S LAST NAME: VOIGTS

DATE/TIME OF TEST: 02/24/2019 14:39:48

Diagnostics Check	OK	14:36	
Air Blank	0.000	14:36	
Control Test	0.078	14:37	
Air Blank	0.000	14:37	
Subject Sample	# 1	14:38	Subject Test Refused
Breath Volume	0.000	14:38	
Air Blank	0.000	14:38	
Control Test	0.078	14:39	
Air Blank	0.000	14:39	
Diagnostics Check	OK	14:39	

Data Download Date/Time: 03/12/2019 08:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HASTINGS,KENNETH,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: HASTINGS,KENNETH

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 02/06/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17817080A3

DRY GAS EXPIRY DATE: 08/05/2019

VIOLATION CODE: DUI

ARREST TIME: 21:15

GENDER: Female

OBSERVATION PERIOD BEGAN: 21:29

SUBJECT'S LAST NAME: TOBAN

DATE/TIME OF TEST: 02/28/2019 21:56:57

Diagnostics Check	OK	21:49
Air Blank	0.000	21:50
Control Test	0.080	21:50
Air Blank	0.000	21:50
Subject Sample # 1	0.095	21:51
Breath Volume	1.332	21:51
Air Blank	0.000	21:52
Air Blank	0.000	21:54
Subject Sample # 2	0.097	21:55
Breath Volume	1.363	21:55
Air Blank	0.000	21:56
Control Test	0.080	21:56
Air Blank	0.000	21:56
Diagnostics Check	OK	21:56

Data Download Date/Time: 03/12/2019 08:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARTNER,ARTHUR,F
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GARTNER,ARTHUR
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: ROCK

DATE/TIME OF TEST: 02/02/2019 00:33:03

Diagnostics Check	OK	00:26
Air Blank	0.000	00:26
Control Test	0.081	00:26
Air Blank	0.000	00:27
Subject Sample # 1	0.171	00:28
Breath Volume	2.160	00:28
Air Blank	0.000	00:28
Air Blank	0.000	00:30
Subject Sample # 2	0.156	00:31
Breath Volume	1.878	00:31
Air Blank	0.000	00:32
Control Test	0.081	00:32
Air Blank	0.000	00:32
Diagnostics Check	OK	00:33

Data Download Date/Time: 02/19/2019 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAUGHERTY,D,M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: DAUGHERTY,D
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: HINSZ

DATE/TIME OF TEST: 02/02/2019 01:11:49

Diagnostics Check	OK	01:05
Air Blank	0.000	01:05
Control Test	0.080	01:06
Air Blank	0.000	01:06
Subject Sample # 1	0.210	01:07
Breath Volume	2.781	01:07
Air Blank	0.000	01:07
Air Blank	0.000	01:09
Subject Sample # 2	0.215	01:10
Breath Volume	2.628	01:10
Air Blank	0.000	01:10
Control Test	0.079	01:11
Air Blank	0.000	01:11
Diagnostics Check	OK	01:11

Data Download Date/Time: 02/19/2019 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANNON, MATTHEW, F
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HANNON, MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 02/02/2019 02:17:52

Diagnostics Check	OK	02:11
Air Blank	0.000	02:12
Control Test	0.081	02:12
Air Blank	0.000	02:13
Subject Sample # 1	0.000	02:13
Breath Volume	3.515	02:13
Air Blank	0.000	02:14
Air Blank	0.000	02:16
Subject Sample # 2	0.000	02:16
Breath Volume	3.250	02:16
Air Blank	0.000	02:16
Control Test	0.080	02:17
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

Data Download Date/Time: 02/19/2019 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,BARRY,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: NIXON,BARRY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 02/02/2019 19:06:46

Diagnostics Check	OK	19:00
Air Blank	0.000	19:01
Control Test	0.079	19:01
Air Blank	0.000	19:02
Subject Sample # 1	0.218	19:02
Breath Volume	1.765	19:02
Air Blank	0.000	19:03
Air Blank	0.000	19:04
Subject Sample # 2	0.225	19:05
Breath Volume	1.988	19:05
Air Blank	0.000	19:05
Control Test	0.079	19:06
Air Blank	0.000	19:06
Diagnostics Check	OK	19:06

Data Download Date/Time: 02/19/2019 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,BARRY,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: NIXON,BARRY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 02/03/2019 03:06:46

Diagnostics Check	OK	03:00
Air Blank	0.000	03:01
Control Test	0.081	03:01
Air Blank	0.000	03:01
Subject Sample # 1	0.194	03:02
Breath Volume	2.621	03:02
Air Blank	0.000	03:03
Air Blank	0.000	03:04
Subject Sample # 2	0.193	03:05
Breath Volume	1.953	03:05
Air Blank	0.000	03:05
Control Test	0.080	03:06
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

Data Download Date/Time: 02/19/2019 16:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COFFEE,JASON,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: CAMPANA,NICK
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: CURRIE

DATE/TIME OF TEST: 02/03/2019 23:47:58

Diagnostics Check	OK	23:41
Air Blank	0.000	23:41
Control Test	0.080	23:42
Air Blank	0.000	23:42
Subject Sample # 1	0.164	23:43
Breath Volume	3.691	23:43
Air Blank	0.000	23:44
Air Blank	0.000	23:45
Subject Sample # 2	0.159	23:46
Breath Volume	2.832	23:46
Air Blank	0.000	23:47
Control Test	0.078	23:47
Air Blank	0.000	23:47
Diagnostics Check	OK	23:47

Data Download Date/Time: 02/19/2019 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUSTOS,ALEJANDRO,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BUSTOS,ALEJANDRO
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: HANLON

DATE/TIME OF TEST: 02/04/2019 03:47:11

Diagnostics Check	OK	03:44	
Air Blank	0.000	03:44	
Control Test	0.080	03:45	
Air Blank	0.000	03:45	
Subject Sample	# 1	03:45	Subject Test Refused
Breath Volume	0.000	03:45	
Air Blank	0.000	03:46	
Control Test	0.080	03:46	
Air Blank	0.000	03:47	
Diagnostics Check	OK	03:47	

Data Download Date/Time: 02/19/2019 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BECKMAN,KEVIN,E
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: BECKMAN,KEVIN
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 13:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:15

SUBJECT'S LAST NAME: CONLEY

DATE/TIME OF TEST: 02/04/2019 14:41:42

Diagnostics Check	OK	14:35
Air Blank	0.000	14:35
Control Test	0.081	14:36
Air Blank	0.000	14:36
Subject Sample # 1	0.000	14:37
Breath Volume	1.410	14:37
Air Blank	0.000	14:37
Air Blank	0.000	14:39
Subject Sample # 2	0.000	14:40
Breath Volume	1.996	14:40
Air Blank	0.000	14:40
Control Test	0.081	14:41
Air Blank	0.000	14:41
Diagnostics Check	OK	14:41

Data Download Date/Time: 02/19/2019 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK, TYLER, J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HAMMELMAN, TYLER
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: BITTINGER

DATE/TIME OF TEST: 02/05/2019 04:47:07

Diagnostics Check	OK	04:40
Air Blank	0.000	04:40
Control Test	0.079	04:41
Air Blank	0.000	04:41
Subject Sample # 1	0.145	04:42
Breath Volume	1.593	04:42
Air Blank	0.000	04:43
Air Blank	0.000	04:45
Subject Sample # 2	0.143	04:45
Breath Volume	1.515	04:45
Air Blank	0.000	04:46
Control Test	0.079	04:46
Air Blank	0.000	04:47
Diagnostics Check	OK	04:47

Data Download Date/Time: 02/19/2019 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARTNER,ARTHUR,F
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GARTNER,ARTHUR
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: BANKAMPEN

DATE/TIME OF TEST: 02/06/2019 00:42:12

Diagnostics Check	OK	00:35
Air Blank	0.000	00:36
Control Test	0.080	00:36
Air Blank	0.000	00:36
Subject Sample # 1	0.000	00:37
Breath Volume	4.960	00:37
Air Blank	0.000	00:38
Air Blank	0.000	00:40
Subject Sample # 2	0.000	00:40
Breath Volume	3.734	00:40
Air Blank	0.000	00:41
Control Test	0.080	00:41
Air Blank	0.000	00:42
Diagnostics Check	OK	00:42

Data Download Date/Time: 02/19/2019 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHANSEN,LLOYD,D
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: REESE,ASHLYN
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 06:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:42

SUBJECT'S LAST NAME: YATES

DATE/TIME OF TEST: 02/07/2019 07:08:52

Diagnostics Check	OK	07:02
Air Blank	0.000	07:03
Control Test	0.080	07:03
Air Blank	0.000	07:03
Subject Sample # 1	0.000	07:04
Breath Volume	3.140	07:04
Air Blank	0.000	07:04
Air Blank	0.000	07:06
Subject Sample # 2	0.000	07:07
Breath Volume	3.667	07:07
Air Blank	0.000	07:07
Control Test	0.079	07:08
Air Blank	0.000	07:08
Diagnostics Check	OK	07:08

Data Download Date/Time: 02/19/2019 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANNON, MATTHEW, F
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HANNON, MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: ASECIO

DATE/TIME OF TEST: 02/08/2019 01:08:01

Diagnostics Check	OK	01:01
Air Blank	0.000	01:02
Control Test	0.081	01:02
Air Blank	0.000	01:02
Subject Sample # 1	0.178	01:03
Breath Volume	2.898	01:03
Air Blank	0.000	01:04
Air Blank	0.000	01:05
Subject Sample # 2	0.185	01:06
Breath Volume	3.078	01:06
Air Blank	0.000	01:07
Control Test	0.080	01:07
Air Blank	0.000	01:07
Diagnostics Check	OK	01:08

Data Download Date/Time: 02/19/2019 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEQUEIRA,NATHAN,C
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SEQUEIRA,NATHAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:05

SUBJECT'S LAST NAME: BILOTTA

DATE/TIME OF TEST: 02/09/2019 03:16:44

Diagnostics Check	OK	03:09
Air Blank	0.000	03:10
Control Test	0.081	03:10
Air Blank	0.000	03:11
Subject Sample # 1	0.086	03:12
Breath Volume	1.582	03:12
Air Blank	0.000	03:12
Air Blank	0.000	03:14
Subject Sample # 2	0.092	03:15
Breath Volume	1.570	03:15
Air Blank	0.000	03:15
Control Test	0.081	03:16
Air Blank	0.000	03:16
Diagnostics Check	OK	03:16

Data Download Date/Time: 02/19/2019 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLY, WESLEY, R
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KELLY, WESLEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 15:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:15

SUBJECT'S LAST NAME: MCCLLOUD

DATE/TIME OF TEST: 02/12/2019 16:45:20

Diagnostics Check	OK	16:37
Air Blank	0.000	16:37
Control Test	0.081	16:37
Air Blank	0.000	16:38
Subject Sample # 1	0.000	16:40
Breath Volume	1.960	16:40
Air Blank	0.000	16:40
Air Blank	0.000	16:42
Subject Sample # 2	0.000	16:43
Breath Volume	1.769	16:43
Air Blank	0.000	16:44
Control Test	0.081	16:44
Air Blank	0.000	16:45
Diagnostics Check	OK	16:45

Data Download Date/Time: 02/19/2019 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,BARRY,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: NIXON,BARRY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: SENGSTOCK

DATE/TIME OF TEST: 02/15/2019 00:14:52

Diagnostics Check	OK	00:08
Air Blank	0.000	00:09
Control Test	0.080	00:09
Air Blank	0.000	00:10
Subject Sample # 1	0.098	00:10
Breath Volume	1.171	00:10
Air Blank	0.000	00:10
Air Blank	0.000	00:12
Subject Sample # 2	0.102	00:13
Breath Volume	1.539	00:13
Air Blank	0.000	00:13
Control Test	0.080	00:14
Air Blank	0.000	00:14
Diagnostics Check	OK	00:14

Data Download Date/Time: 02/19/2019 16:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOTO,SHANI,R
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: RUNNER,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: SOTO

DATE/TIME OF TEST: 02/15/2019 00:59:58

Diagnostics Check	OK	00:53
Air Blank	0.000	00:53
Control Test	0.079	00:54
Air Blank	0.000	00:54
Subject Sample # 1	0.113	00:55
Breath Volume	1.500	00:55
Air Blank	0.000	00:56
Air Blank	0.000	00:58
Subject Sample # 2	0.105	00:58
Breath Volume	1.109	00:58
Air Blank	0.000	00:59
Control Test	0.079	00:59
Air Blank	0.000	00:59
Diagnostics Check	OK	00:59

Data Download Date/Time: 02/19/2019 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAUGHERTY,D,M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: DAUGHERTY,D
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: PILARTE

DATE/TIME OF TEST: 02/16/2019 00:56:32

Diagnostics Check	OK	00:48
Air Blank	0.000	00:48
Control Test	0.081	00:49
Air Blank	0.000	00:49
Subject Sample # 1	0.079	00:51
Breath Volume	1.167	00:51
Air Blank	0.000	00:52
Air Blank	0.000	00:53
Subject Sample # 2	0.073	00:55
Breath Volume	1.578	00:55
Air Blank	0.000	00:55
Control Test	0.081	00:55
Air Blank	0.000	00:56
Diagnostics Check	OK	00:56

Data Download Date/Time: 02/19/2019 16:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KJEKSTAD,ERIC,C
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KJEKSTAD,ERIC
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: MCCULLOUGH

DATE/TIME OF TEST: 02/16/2019 02:54:12

Diagnostics Check	OK	02:40	
Air Blank	0.000	02:41	
Control Test	0.080	02:41	
Air Blank	0.000	02:41	
Subject Sample	# 1 0.191	02:45	Volume Not Met
Breath Volume	0.425	02:45	
Air Blank	0.000	02:46	
Air Blank	0.000	02:47	
Subject Sample	# 2 0.204	02:49	
Breath Volume	2.082	02:49	
Air Blank	0.000	02:50	
Air Blank	0.000	02:52	
Subject Sample	# 3 0.207	02:52	
Breath Volume	2.734	02:52	
Air Blank	0.000	02:53	
Control Test	0.078	02:53	
Air Blank	0.000	02:54	
Diagnostics Check	OK	02:54	

Data Download Date/Time: 02/19/2019 16:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAUGHERTY,D,M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: DAUGHERTY,D
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: FORD

DATE/TIME OF TEST: 02/17/2019 02:53:17

Diagnostics Check	OK	02:50	
Air Blank	0.000	02:50	
Control Test	0.081	02:51	
Air Blank	0.000	02:51	
Subject Sample	# 1	02:51	Subject Test Refused
Breath Volume	0.000	02:51	
Air Blank	0.000	02:52	
Control Test	0.080	02:52	
Air Blank	0.000	02:53	
Diagnostics Check	OK	02:53	

Data Download Date/Time: 02/19/2019 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANNON, MATTHEW, F
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HANNON, MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:03

SUBJECT'S LAST NAME: KLINE

DATE/TIME OF TEST: 02/17/2019 22:01:50

Diagnostics Check	OK	21:52	
Air Blank	0.000	21:53	
Control Test	0.080	21:53	
Air Blank	0.000	21:53	
Subject Sample	# 1	21:56	No Sample Provided
Breath Volume	0.000	21:56	
Air Blank	0.000	21:57	
Air Blank	0.000	21:59	
Subject Sample	# 2	22:00	Subject Test Refused
Breath Volume	0.000	22:00	
Air Blank	0.000	22:00	
Control Test	0.081	22:01	
Air Blank	0.000	22:01	
Diagnostics Check	OK	22:01	

Data Download Date/Time: 02/19/2019 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUSTOS,ALEJANDRO,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BUSTOS,ALEJANDRO
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 02/18/2019 00:36:32

Diagnostics Check	OK	00:30
Air Blank	0.000	00:30
Control Test	0.080	00:30
Air Blank	0.000	00:31
Subject Sample # 1	0.236	00:32
Breath Volume	2.988	00:32
Air Blank	0.000	00:32
Air Blank	0.000	00:34
Subject Sample # 2	0.233	00:35
Breath Volume	2.503	00:35
Air Blank	0.000	00:35
Control Test	0.079	00:35
Air Blank	0.000	00:36
Diagnostics Check	OK	00:36

Data Download Date/Time: 02/19/2019 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLY, WESLEY, R
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KELLY, WESLEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:55

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 02/20/2019 18:21:57

Diagnostics Check	OK	18:15
Air Blank	0.000	18:15
Control Test	0.080	18:16
Air Blank	0.000	18:16
Subject Sample # 1	0.277	18:17
Breath Volume	1.925	18:17
Air Blank	0.000	18:17
Air Blank	0.000	18:19
Subject Sample # 2	0.285	18:20
Breath Volume	2.316	18:20
Air Blank	0.000	18:21
Control Test	0.079	18:21
Air Blank	0.000	18:21
Diagnostics Check	OK	18:21

Data Download Date/Time: 03/29/2019 22:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,BARRY,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: NIXON,BARRY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 02/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: DODICH

DATE/TIME OF TEST: 02/20/2019 21:11:34

Diagnostics Check	OK	21:05
Air Blank	0.000	21:06
Control Test	0.079	21:06
Air Blank	0.000	21:06
Subject Sample # 1	0.226	21:07
Breath Volume	1.707	21:07
Air Blank	0.000	21:07
Air Blank	0.000	21:09
Subject Sample # 2	0.230	21:10
Breath Volume	1.765	21:10
Air Blank	0.000	21:10
Control Test	0.079	21:10
Air Blank	0.000	21:11
Diagnostics Check	OK	21:11

Data Download Date/Time: 03/29/2019 22:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLY, WESLEY, R
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KELLY, WESLEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 15:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: FRASER

DATE/TIME OF TEST: 02/21/2019 17:42:23

Diagnostics Check	OK	17:35
Air Blank	0.000	17:36
Control Test	0.080	17:36
Air Blank	0.000	17:36
Subject Sample # 1	0.168	17:37
Breath Volume	3.066	17:37
Air Blank	0.000	17:38
Air Blank	0.000	17:40
Subject Sample # 2	0.162	17:40
Breath Volume	2.589	17:40
Air Blank	0.000	17:41
Control Test	0.080	17:41
Air Blank	0.000	17:42
Diagnostics Check	OK	17:42

Data Download Date/Time: 03/29/2019 22:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUSTOS,ALEJANDRO,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BUSTOS,ALEJANDRO
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:48

SUBJECT'S LAST NAME: DICKSON

DATE/TIME OF TEST: 02/21/2019 22:13:24

Diagnostics Check	OK	22:07	
Air Blank	0.000	22:08	
Control Test	0.080	22:08	
Air Blank	0.000	22:09	
Subject Sample	# 1	22:12	Subject Test Refused
Breath Volume	0.000	22:12	
Air Blank	0.000	22:12	
Control Test	0.080	22:12	
Air Blank	0.000	22:13	
Diagnostics Check	OK	22:13	

Data Download Date/Time: 03/29/2019 22:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,BARRY,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: NIXON,BARRY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 02/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:03

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 02/23/2019 22:29:58

Diagnostics Check	OK	22:23
Air Blank	0.000	22:23
Control Test	0.081	22:23
Air Blank	0.000	22:24
Subject Sample # 1	0.219	22:24
Breath Volume	1.335	22:24
Air Blank	0.000	22:25
Air Blank	0.000	22:27
Subject Sample # 2	0.200	22:28
Breath Volume	2.394	22:28
Air Blank	0.000	22:29
Control Test	0.081	22:29
Air Blank	0.000	22:29
Diagnostics Check	OK	22:29

Data Download Date/Time: 03/29/2019 22:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK, TYLER, J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KIRK, TYLER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: B200721930970

DATE/TIME OF TEST: 02/23/2019 23:53:04

Diagnostics Check	OK	23:50	
Air Blank	0.000	23:50	
Control Test	0.079	23:51	
Air Blank	0.000	23:51	
Subject Sample	# 1	23:51	Subject Test Refused
Breath Volume	0.000	23:51	
Air Blank	0.000	23:52	
Control Test	0.080	23:52	
Air Blank	0.000	23:52	
Diagnostics Check	OK	23:53	

Data Download Date/Time: 03/29/2019 22:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,RAEFORD,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GRIFFIN,RAEFORD
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: WEICHT

DATE/TIME OF TEST: 02/24/2019 04:08:18

Diagnostics Check	OK	04:01
Air Blank	0.000	04:02
Control Test	0.079	04:02
Air Blank	0.000	04:03
Subject Sample # 1	0.168	04:03
Breath Volume	3.734	04:03
Air Blank	0.000	04:04
Air Blank	0.000	04:06
Subject Sample # 2	0.165	04:06
Breath Volume	3.660	04:06
Air Blank	0.000	04:07
Control Test	0.079	04:07
Air Blank	0.000	04:08
Diagnostics Check	OK	04:08

Data Download Date/Time: 03/29/2019 22:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REAY,ALEX,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: REAY,ALEX
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 02/25/2019 18:40:00

Diagnostics Check	OK	18:37	
Air Blank	0.000	18:37	
Control Test	0.082	18:37	
Air Blank	0.000	18:38	
Subject Sample	# 1	18:38	Subject Test Refused
Breath Volume	0.000	18:38	
Air Blank	0.000	18:39	
Control Test	0.081	18:39	
Air Blank	0.000	18:39	
Diagnostics Check	OK	18:40	

Data Download Date/Time: 03/29/2019 22:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANNON, MATTHEW, F
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HANNON, MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: LIEBERSTEIN

DATE/TIME OF TEST: 02/27/2019 00:41:34

Diagnostics Check	OK	00:35
Air Blank	0.000	00:35
Control Test	0.081	00:35
Air Blank	0.000	00:36
Subject Sample # 1	0.173	00:37
Breath Volume	3.691	00:37
Air Blank	0.000	00:37
Air Blank	0.000	00:39
Subject Sample # 2	0.169	00:40
Breath Volume	3.441	00:40
Air Blank	0.000	00:40
Control Test	0.080	00:40
Air Blank	0.000	00:41
Diagnostics Check	OK	00:41

Data Download Date/Time: 03/29/2019 22:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUSTOS,ALEJANDRO,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BUSTOS,ALEJANDRO
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: HEYE

DATE/TIME OF TEST: 02/27/2019 04:48:14

Diagnostics Check	OK	04:41
Air Blank	0.000	04:41
Control Test	0.081	04:41
Air Blank	0.000	04:42
Subject Sample # 1	0.043	04:43
Breath Volume	1.871	04:43
Air Blank	0.000	04:44
Air Blank	0.000	04:46
Subject Sample # 2	0.040	04:46
Breath Volume	1.929	04:46
Air Blank	0.000	04:47
Control Test	0.079	04:47
Air Blank	0.000	04:48
Diagnostics Check	OK	04:48

Data Download Date/Time: 03/29/2019 22:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLIS JR,MICHAEL,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HOLLIS JR,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: GRAY

DATE/TIME OF TEST: 02/28/2019 02:41:20

Diagnostics Check	OK	02:35
Air Blank	0.000	02:35
Control Test	0.080	02:35
Air Blank	0.000	02:36
Subject Sample # 1	0.083	02:37
Breath Volume	3.316	02:37
Air Blank	0.000	02:37
Air Blank	0.000	02:39
Subject Sample # 2	0.067	02:39
Breath Volume	1.519	02:39
Air Blank	0.000	02:40
Control Test	0.080	02:40
Air Blank	0.000	02:41
Diagnostics Check	OK	02:41

Data Download Date/Time: 03/29/2019 22:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Port Richey PD
ARRESTING OFFICER: WILLIAMS,ROBERT
ARRESTING OFFICER AGENCY: Port Richey PD
LAST AGENCY INSPECTION: 02/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 15:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: LOOMIS

DATE/TIME OF TEST: 02/28/2019 17:25:20

Diagnostics Check	OK	17:10	
Air Blank	0.000	17:11	
Control Test	0.080	17:11	
Air Blank	0.000	17:12	
Subject Sample # 1	0.192	17:14	
Breath Volume	1.125	17:14	
Air Blank	0.000	17:15	
Air Blank	0.000	17:17	
Subject Sample # 2	0.160	17:20	Volume Not Met
Breath Volume	0.117	17:20	
Air Blank	0.000	17:21	
Air Blank	0.000	17:22	
Subject Sample # 3	0.190	17:23	
Breath Volume	1.500	17:23	
Air Blank	0.000	17:24	
Control Test	0.078	17:24	
Air Blank	0.000	17:25	
Diagnostics Check	OK	17:25	

Data Download Date/Time: 03/29/2019 22:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARLEY,OFC K,G
BREATH TEST OPERATOR AGENCY: Plantation PD
ARRESTING OFFICER: FARQUHARSON,OFC K
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 01/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A5
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: DUHANEY

DATE/TIME OF TEST: 02/14/2019 01:51:57

Diagnostics Check	OK	01:49	
Air Blank	0.000	01:49	
Control Test	0.078	01:49	
Air Blank	0.000	01:50	
Subject Sample	# 1	01:50	Subject Test Refused
Breath Volume	0.000	01:50	
Air Blank	0.000	01:51	
Control Test	0.078	01:51	
Air Blank	0.000	01:51	
Diagnostics Check	OK	01:51	

Data Download Date/Time: 02/27/2019 20:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARQUHARSON,KEVIN,D
BREATH TEST OPERATOR AGENCY: Plantation PD
ARRESTING OFFICER: BUKATA,AMANDA
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 01/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A5
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 05:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:19

SUBJECT'S LAST NAME: RUIZ

DATE/TIME OF TEST: 02/14/2019 06:46:58

Diagnostics Check	OK	06:40
Air Blank	0.000	06:40
Control Test	0.079	06:41
Air Blank	0.000	06:41
Subject Sample # 1	0.178	06:42
Breath Volume	1.859	06:42
Air Blank	0.000	06:42
Air Blank	0.000	06:44
Subject Sample # 2	0.171	06:45
Breath Volume	1.488	06:45
Air Blank	0.000	06:46
Control Test	0.078	06:46
Air Blank	0.000	06:46
Diagnostics Check	OK	06:46

Data Download Date/Time: 02/27/2019 20:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROSKIE,OFC J,
BREATH TEST OPERATOR AGENCY: Plantation PD
ARRESTING OFFICER: MULLINGS,OFC B
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 01/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A5
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: DOWNER

DATE/TIME OF TEST: 02/15/2019 21:26:39

Diagnostics Check	OK	21:20
Air Blank	0.000	21:21
Control Test	0.078	21:21
Air Blank	0.000	21:21
Subject Sample # 1	0.243	21:22
Breath Volume	2.121	21:22
Air Blank	0.000	21:22
Air Blank	0.000	21:24
Subject Sample # 2	0.253	21:25
Breath Volume	1.957	21:25
Air Blank	0.000	21:25
Control Test	0.075	21:26
Air Blank	0.000	21:26
Diagnostics Check	OK	21:26

Data Download Date/Time: 02/27/2019 20:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARQUHARSON,KEVIN,D
BREATH TEST OPERATOR AGENCY: Plantation PD
ARRESTING OFFICER: FARQUHARSON,KEVIN
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 01/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A5
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: FAVROT

DATE/TIME OF TEST: 02/17/2019 04:38:51

Diagnostics Check	OK	04:32
Air Blank	0.000	04:32
Control Test	0.078	04:32
Air Blank	0.000	04:33
Subject Sample # 1	0.053	04:34
Breath Volume	3.207	04:34
Air Blank	0.000	04:34
Air Blank	0.000	04:36
Subject Sample # 2	0.050	04:37
Breath Volume	3.398	04:37
Air Blank	0.000	04:37
Control Test	0.078	04:38
Air Blank	0.000	04:38
Diagnostics Check	OK	04:38

Data Download Date/Time: 02/27/2019 20:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARQUHARSON,KEVIN,D
BREATH TEST OPERATOR AGENCY: Plantation PD
ARRESTING OFFICER: FARQUHARSON,KEVIN
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 01/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A5
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: PAUL

DATE/TIME OF TEST: 02/21/2019 23:09:18

Diagnostics Check	OK	23:02
Air Blank	0.000	23:02
Control Test	0.078	23:03
Air Blank	0.000	23:03
Subject Sample # 1	0.254	23:04
Breath Volume	1.457	23:04
Air Blank	0.000	23:04
Air Blank	0.000	23:06
Subject Sample # 2	0.240	23:07
Breath Volume	1.289	23:07
Air Blank	0.000	23:08
Control Test	0.078	23:08
Air Blank	0.000	23:09
Diagnostics Check	OK	23:09

Data Download Date/Time: 02/27/2019 20:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARQUHARSON,KEVIN,D
BREATH TEST OPERATOR AGENCY: Plantation PD
ARRESTING OFFICER: FARQUHARSON,KEVIN
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 01/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A5
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: BENJAMIN

DATE/TIME OF TEST: 02/23/2019 05:12:24

Diagnostics Check	OK	05:05
Air Blank	0.000	05:05
Control Test	0.078	05:06
Air Blank	0.000	05:06
Subject Sample # 1	0.145	05:07
Breath Volume	1.835	05:07
Air Blank	0.000	05:08
Air Blank	0.000	05:10
Subject Sample # 2	0.143	05:10
Breath Volume	1.710	05:10
Air Blank	0.000	05:11
Control Test	0.078	05:11
Air Blank	0.000	05:12
Diagnostics Check	OK	05:12

Data Download Date/Time: 02/27/2019 20:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: STREATER

DATE/TIME OF TEST: 02/01/2019 00:28:17

Diagnostics Check	OK	00:22
Air Blank	0.000	00:22
Control Test	0.081	00:22
Air Blank	0.000	00:23
Subject Sample # 1	0.132	00:23
Breath Volume	3.371	00:23
Air Blank	0.000	00:24
Air Blank	0.000	00:26
Subject Sample # 2	0.127	00:26
Breath Volume	2.761	00:26
Air Blank	0.000	00:27
Control Test	0.081	00:27
Air Blank	0.000	00:28
Diagnostics Check	OK	00:28

Data Download Date/Time: 02/07/2019 13:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: WOLF

DATE/TIME OF TEST: 02/01/2019 03:54:42

Diagnostics Check	OK	03:51	
Air Blank	0.000	03:52	
Control Test	0.081	03:52	
Air Blank	0.000	03:53	
Subject Sample	# 1	03:53	Subject Test Refused
Breath Volume	0.000	03:53	
Air Blank	0.000	03:53	
Control Test	0.080	03:54	
Air Blank	0.000	03:54	
Diagnostics Check	OK	03:54	

Data Download Date/Time: 02/07/2019 13:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,D,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 02/02/2019 00:01:52

Diagnostics Check	OK	23:55
Air Blank	0.000	23:55
Control Test	0.081	23:56
Air Blank	0.000	23:56
Subject Sample # 2	0.143	23:57
Breath Volume	1.734	23:57
Air Blank	0.000	23:57
Air Blank	0.000	23:59
Subject Sample # 1	0.144	00:00
Breath Volume	1.941	00:00
Air Blank	0.000	00:00
Control Test	0.081	00:01
Air Blank	0.000	00:01
Diagnostics Check	OK	00:01

Data Download Date/Time: 02/07/2019 13:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,D,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: BODIN

DATE/TIME OF TEST: 02/02/2019 02:08:28

Diagnostics Check	OK	02:05	
Air Blank	0.000	02:06	
Control Test	0.081	02:06	
Air Blank	0.000	02:06	
Subject Sample	# 1	02:07	Subject Test Refused
Breath Volume	0.000	02:07	
Air Blank	0.000	02:07	
Control Test	0.081	02:07	
Air Blank	0.000	02:08	
Diagnostics Check	OK	02:08	

Data Download Date/Time: 02/07/2019 13:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,D,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: POTTER

DATE/TIME OF TEST: 02/02/2019 04:27:37

Diagnostics Check	OK	04:21
Air Blank	0.000	04:21
Control Test	0.081	04:22
Air Blank	0.000	04:22
Subject Sample # 1	0.111	04:23
Breath Volume	4.433	04:23
Air Blank	0.000	04:23
Air Blank	0.000	04:25
Subject Sample # 2	0.109	04:26
Breath Volume	4.464	04:26
Air Blank	0.000	04:26
Control Test	0.081	04:27
Air Blank	0.000	04:27
Diagnostics Check	OK	04:27

Data Download Date/Time: 02/07/2019 13:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,D,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: FROST

DATE/TIME OF TEST: 02/04/2019 00:31:22

Diagnostics Check	OK	00:28	
Air Blank	0.000	00:28	
Control Test	0.080	00:29	
Air Blank	0.000	00:29	
Subject Sample	# 1	00:29	Subject Test Refused
Breath Volume	0.000	00:29	
Air Blank	0.000	00:30	
Control Test	0.080	00:30	
Air Blank	0.000	00:31	
Diagnostics Check	OK	00:31	

Data Download Date/Time: 02/07/2019 13:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,D,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FILBERT,E
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: LIVINGSTON

DATE/TIME OF TEST: 02/04/2019 01:13:37

Diagnostics Check	OK	01:10	
Air Blank	0.000	01:11	
Control Test	0.080	01:11	
Air Blank	0.000	01:11	
Subject Sample	# 1	01:12	Subject Test Refused
Breath Volume	0.000	01:12	
Air Blank	0.000	01:12	
Control Test	0.080	01:13	
Air Blank	0.000	01:13	
Diagnostics Check	OK	01:13	

Data Download Date/Time: 02/07/2019 13:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,D,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: PETERS

DATE/TIME OF TEST: 02/04/2019 03:04:26

Diagnostics Check	OK	03:01	
Air Blank	0.000	03:02	
Control Test	0.080	03:02	
Air Blank	0.000	03:02	
Subject Sample	# 1	03:03	Subject Test Refused
Breath Volume	0.000	03:03	
Air Blank	0.000	03:03	
Control Test	0.081	03:03	
Air Blank	0.000	03:04	
Diagnostics Check	OK	03:04	

Data Download Date/Time: 02/07/2019 13:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,D,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SORRELLS,M
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: REGALADO MONROY

DATE/TIME OF TEST: 02/04/2019 04:07:00

Diagnostics Check	OK	03:59
Air Blank	0.000	04:00
Control Test	0.080	04:00
Air Blank	0.000	04:01
Subject Sample # 1	0.125	04:02
Breath Volume	3.074	04:02
Air Blank	0.000	04:03
Air Blank	0.000	04:04
Subject Sample # 2	0.129	04:05
Breath Volume	3.496	04:05
Air Blank	0.000	04:06
Control Test	0.080	04:06
Air Blank	0.000	04:06
Diagnostics Check	OK	04:07

Data Download Date/Time: 02/07/2019 13:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,D,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: ST CLAIR

DATE/TIME OF TEST: 02/04/2019 23:21:06

Diagnostics Check	OK	23:14
Air Blank	0.000	23:14
Control Test	0.081	23:15
Air Blank	0.000	23:15
Subject Sample # 1	0.137	23:16
Breath Volume	1.976	23:16
Air Blank	0.000	23:17
Air Blank	0.000	23:18
Subject Sample # 2	0.133	23:19
Breath Volume	2.847	23:19
Air Blank	0.000	23:20
Control Test	0.080	23:20
Air Blank	0.000	23:21
Diagnostics Check	OK	23:21

Data Download Date/Time: 02/07/2019 13:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,M,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRYANT,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: GLASPIE

DATE/TIME OF TEST: 02/08/2019 03:56:53

Diagnostics Check	OK	03:53	
Air Blank	0.000	03:54	
Control Test	0.081	03:54	
Air Blank	0.000	03:55	
Subject Sample	# 1	03:55	Subject Test Refused
Breath Volume	0.000	03:55	
Air Blank	0.000	03:55	
Control Test	0.082	03:56	
Air Blank	0.000	03:56	
Diagnostics Check	OK	03:56	

Data Download Date/Time: 03/04/2019 13:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AUGUSTINE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: MCKENZY

DATE/TIME OF TEST: 02/09/2019 01:06:10

Diagnostics Check	OK	01:03	
Air Blank	0.000	01:03	
Control Test	0.081	01:04	
Air Blank	0.000	01:04	
Subject Sample	# 1	01:04	Subject Test Refused
Breath Volume	0.000	01:04	
Air Blank	0.000	01:05	
Control Test	0.081	01:05	
Air Blank	0.000	01:06	
Diagnostics Check	OK	01:06	

Data Download Date/Time: 03/04/2019 13:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AUGUSTINE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 02/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: KEENE

DATE/TIME OF TEST: 02/10/2019 03:31:48

Diagnostics Check	OK	03:27	
Air Blank	0.000	03:27	
Control Test	0.081	03:28	
Air Blank	0.000	03:28	
Subject Sample	# 1	03:30	Subject Test Refused
Breath Volume	0.000	03:30	
Air Blank	0.000	03:30	
Control Test	0.081	03:31	
Air Blank	0.000	03:31	
Diagnostics Check	OK	03:31	

Data Download Date/Time: 03/04/2019 13:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AUGUSTINE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WILLIAMSON,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 02/10/2019 04:41:13

Diagnostics Check	OK	04:30	
Air Blank	0.000	04:31	
Control Test	0.081	04:31	
Air Blank	0.000	04:32	
Subject Sample	# 1 0.178	04:35	Volume Not Met
Breath Volume	0.050	04:35	
Air Blank	0.000	04:36	
Air Blank	0.000	04:37	
Subject Sample	# 2	04:39	Subject Test Refused
Breath Volume	0.109	04:39	
Air Blank	0.000	04:40	
Control Test	0.079	04:40	
Air Blank	0.000	04:41	
Diagnostics Check	OK	04:41	

Data Download Date/Time: 03/04/2019 13:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AUGUSTINE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 02/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: HARRISON

DATE/TIME OF TEST: 02/11/2019 03:44:37

Diagnostics Check	OK	03:41	
Air Blank	0.000	03:42	
Control Test	0.081	03:42	
Air Blank	0.000	03:42	
Subject Sample	# 1	03:43	Subject Test Refused
Breath Volume	0.000	03:43	
Air Blank	0.000	03:43	
Control Test	0.081	03:44	
Air Blank	0.000	03:44	
Diagnostics Check	OK	03:44	

Data Download Date/Time: 03/04/2019 13:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,D,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMS,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: SWINEFORD

DATE/TIME OF TEST: 02/14/2019 00:59:16

Diagnostics Check	OK	00:54	
Air Blank	0.000	00:55	
Control Test	0.081	00:55	
Air Blank	0.000	00:56	
Subject Sample	# 1	00:57	Subject Test Refused
Breath Volume	0.000	00:57	
Air Blank	0.000	00:58	
Control Test	0.082	00:58	
Air Blank	0.000	00:59	
Diagnostics Check	OK	00:59	

Data Download Date/Time: 03/04/2019 13:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,D,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: PARRISH

DATE/TIME OF TEST: 02/14/2019 01:57:34

Diagnostics Check	OK	01:51
Air Blank	0.000	01:51
Control Test	0.082	01:51
Air Blank	0.000	01:52
Subject Sample # 1	0.081	01:52
Breath Volume	1.859	01:52
Air Blank	0.000	01:53
Air Blank	0.000	01:55
Subject Sample # 2	0.081	01:56
Breath Volume	2.195	01:56
Air Blank	0.000	01:56
Control Test	0.081	01:56
Air Blank	0.000	01:57
Diagnostics Check	OK	01:57

Data Download Date/Time: 03/04/2019 13:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,D,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: CANSECO RAMIREZ

DATE/TIME OF TEST: 02/14/2019 03:26:04

Diagnostics Check	OK	03:19
Air Blank	0.000	03:19
Control Test	0.081	03:19
Air Blank	0.000	03:20
Subject Sample # 1	0.086	03:21
Breath Volume	1.617	03:21
Air Blank	0.000	03:21
Air Blank	0.000	03:23
Subject Sample # 2	0.083	03:24
Breath Volume	1.226	03:24
Air Blank	0.000	03:25
Control Test	0.081	03:25
Air Blank	0.000	03:25
Diagnostics Check	OK	03:26

Data Download Date/Time: 03/04/2019 13:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AUGUSTINE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: GRAHAM

DATE/TIME OF TEST: 02/17/2019 00:06:07

Diagnostics Check	OK	00:03	
Air Blank	0.000	00:03	
Control Test	0.080	00:03	
Air Blank	0.000	00:04	
Subject Sample	# 1	00:04	Subject Test Refused
Breath Volume	0.000	00:04	
Air Blank	0.000	00:05	
Control Test	0.081	00:05	
Air Blank	0.000	00:06	
Diagnostics Check	OK	00:06	

Data Download Date/Time: 03/04/2019 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AUGUSTINE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 02/17/2019 01:24:15

Diagnostics Check	OK	01:17
Air Blank	0.000	01:17
Control Test	0.080	01:18
Air Blank	0.000	01:18
Subject Sample # 1	0.198	01:19
Breath Volume	1.746	01:19
Air Blank	0.000	01:20
Air Blank	0.000	01:22
Subject Sample # 2	0.199	01:22
Breath Volume	1.644	01:22
Air Blank	0.000	01:23
Control Test	0.080	01:23
Air Blank	0.000	01:24
Diagnostics Check	OK	01:24

Data Download Date/Time: 03/04/2019 13:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AUGUSTINE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: MAGDON

DATE/TIME OF TEST: 02/17/2019 04:06:08

Diagnostics Check	OK	04:03	
Air Blank	0.000	04:03	
Control Test	0.081	04:04	
Air Blank	0.000	04:04	
Subject Sample	# 1	04:04	Subject Test Refused
Breath Volume	0.000	04:04	
Air Blank	0.000	04:05	
Control Test	0.081	04:05	
Air Blank	0.000	04:06	
Diagnostics Check	OK	04:06	

Data Download Date/Time: 03/04/2019 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AUGUSTINE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HAVENS,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:16

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 02/17/2019 05:42:44

Diagnostics Check	OK	05:36
Air Blank	0.000	05:36
Control Test	0.081	05:36
Air Blank	0.000	05:37
Subject Sample # 1	0.123	05:37
Breath Volume	1.671	05:37
Air Blank	0.000	05:38
Air Blank	0.000	05:40
Subject Sample # 2	0.127	05:41
Breath Volume	1.464	05:41
Air Blank	0.000	05:41
Control Test	0.080	05:42
Air Blank	0.000	05:42
Diagnostics Check	OK	05:42

Data Download Date/Time: 03/04/2019 13:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAMES,A,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: OLIVERAS,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 05:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:48

SUBJECT'S LAST NAME: LOGAN

DATE/TIME OF TEST: 02/17/2019 08:16:43

Diagnostics Check	OK	08:09
Air Blank	0.000	08:09
Control Test	0.079	08:10
Air Blank	0.000	08:10
Subject Sample # 1	0.136	08:12
Breath Volume	2.164	08:12
Air Blank	0.000	08:12
Air Blank	0.000	08:14
Subject Sample # 2	0.141	08:15
Breath Volume	2.277	08:15
Air Blank	0.000	08:15
Control Test	0.079	08:16
Air Blank	0.000	08:16
Diagnostics Check	OK	08:16

Data Download Date/Time: 03/04/2019 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AUGUSTINE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PARMENTER,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: HERIBERTO

DATE/TIME OF TEST: 02/27/2019 23:52:07

Diagnostics Check	OK	23:45
Air Blank	0.000	23:46
Control Test	0.080	23:46
Air Blank	0.000	23:46
Subject Sample # 1	0.154	23:47
Breath Volume	2.757	23:47
Air Blank	0.000	23:48
Air Blank	0.000	23:49
Subject Sample # 2	0.145	23:50
Breath Volume	2.828	23:50
Air Blank	0.000	23:51
Control Test	0.080	23:51
Air Blank	0.000	23:52
Diagnostics Check	OK	23:52

Data Download Date/Time: 03/04/2019 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AUGUSTINE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MILLER,T
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: CLEMENTS

DATE/TIME OF TEST: 02/28/2019 03:07:15

Diagnostics Check	OK	02:59
Air Blank	0.000	02:59
Control Test	0.080	03:00
Air Blank	0.000	03:00
Subject Sample # 1	0.134	03:01
Breath Volume	2.000	03:01
Air Blank	0.000	03:02
Air Blank	0.000	03:03
Subject Sample # 2	0.135	03:05
Breath Volume	1.101	03:05
Air Blank	0.000	03:06
Control Test	0.080	03:06
Air Blank	0.000	03:07
Diagnostics Check	OK	03:07

Data Download Date/Time: 03/04/2019 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AUGUSTINE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: PAIGE

DATE/TIME OF TEST: 02/28/2019 03:54:11

Diagnostics Check	OK	03:44	
Air Blank	0.000	03:44	
Control Test	0.079	03:45	
Air Blank	0.000	03:45	
Subject Sample	# 1 0.087	03:48	Volume Not Met
Breath Volume	1.089	03:48	
Air Blank	0.000	03:49	
Air Blank	0.000	03:51	
Subject Sample	# 2	03:52	Subject Test Refused
Breath Volume	0.015	03:52	
Air Blank	0.000	03:53	
Control Test	0.079	03:53	
Air Blank	0.000	03:54	
Diagnostics Check	OK	03:54	

Data Download Date/Time: 03/04/2019 13:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AUGUSTINE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNS,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 02/28/2019 05:07:49

Diagnostics Check	OK	05:04	
Air Blank	0.000	05:05	
Control Test	0.081	05:05	
Air Blank	0.000	05:06	
Subject Sample	# 1	05:06	Subject Test Refused
Breath Volume	0.000	05:06	
Air Blank	0.000	05:06	
Control Test	0.080	05:07	
Air Blank	0.000	05:07	
Diagnostics Check	OK	05:07	

Data Download Date/Time: 03/04/2019 13:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWMAN,JOANNA,N
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BOWMAN,JOANNA
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: WICKMAN

DATE/TIME OF TEST: 02/05/2019 04:42:39

Diagnostics Check	OK	04:35
Air Blank	0.000	04:35
Control Test	0.081	04:36
Air Blank	0.000	04:36
Subject Sample # 1	0.162	04:38
Breath Volume	2.042	04:38
Air Blank	0.000	04:38
Air Blank	0.000	04:40
Subject Sample # 2	0.153	04:41
Breath Volume	1.851	04:41
Air Blank	0.000	04:41
Control Test	0.081	04:42
Air Blank	0.000	04:42
Diagnostics Check	OK	04:42

Data Download Date/Time: 02/14/2019 16:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,RYAN,W
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BROWN,RYAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 16:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:01

SUBJECT'S LAST NAME: REDMOND

DATE/TIME OF TEST: 02/12/2019 17:41:23

Diagnostics Check	OK	17:35
Air Blank	0.000	17:35
Control Test	0.080	17:36
Air Blank	0.000	17:36
Subject Sample # 1	0.000	17:37
Breath Volume	2.351	17:37
Air Blank	0.000	17:37
Air Blank	0.000	17:39
Subject Sample # 2	0.000	17:40
Breath Volume	2.136	17:40
Air Blank	0.000	17:40
Control Test	0.081	17:40
Air Blank	0.000	17:41
Diagnostics Check	OK	17:41

Data Download Date/Time: 02/14/2019 16:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,JACOB,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MOORE,JACOB
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: WILKINS

DATE/TIME OF TEST: 02/20/2019 03:01:55

Diagnostics Check	OK	02:54
Air Blank	0.000	02:54
Control Test	0.080	02:55
Air Blank	0.000	02:55
Subject Sample # 1	0.178	02:57
Breath Volume	2.363	02:57
Air Blank	0.000	02:57
Air Blank	0.000	02:59
Subject Sample # 2	0.178	03:00
Breath Volume	2.210	03:00
Air Blank	0.000	03:01
Control Test	0.080	03:01
Air Blank	0.000	03:01
Diagnostics Check	OK	03:01

Data Download Date/Time: 03/18/2019 17:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GENTRY,JONATHAN,P
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: GENTRY,JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: ROBERTSON

DATE/TIME OF TEST: 02/23/2019 01:17:25

Diagnostics Check	OK	01:10
Air Blank	0.000	01:11
Control Test	0.081	01:11
Air Blank	0.000	01:11
Subject Sample # 1	0.204	01:12
Breath Volume	2.812	01:12
Air Blank	0.000	01:13
Air Blank	0.000	01:15
Subject Sample # 2	0.213	01:15
Breath Volume	2.597	01:15
Air Blank	0.000	01:16
Control Test	0.081	01:16
Air Blank	0.000	01:17
Diagnostics Check	OK	01:17

Data Download Date/Time: 03/18/2019 17:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISK,CARLOS,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: SISK,CARLOS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/13/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: FULLER

DATE/TIME OF TEST: 02/23/2019 02:06:24

Diagnostics Check	OK	02:03	
Air Blank	0.000	02:03	
Control Test	0.081	02:04	
Air Blank	0.000	02:04	
Subject Sample	# 1	02:05	Subject Test Refused
Breath Volume	0.000	02:05	
Air Blank	0.000	02:05	
Control Test	0.080	02:05	
Air Blank	0.000	02:06	
Diagnostics Check	OK	02:06	

Data Download Date/Time: 03/18/2019 17:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISK,CARLOS,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: SISK,CARLOS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: PUCKETT

DATE/TIME OF TEST: 02/24/2019 03:45:54

Diagnostics Check	OK	03:43	
Air Blank	0.000	03:43	
Control Test	0.081	03:43	
Air Blank	0.000	03:44	
Subject Sample	# 1	03:44	Subject Test Refused
Breath Volume	0.000	03:44	
Air Blank	0.000	03:45	
Control Test	0.081	03:45	
Air Blank	0.000	03:45	
Diagnostics Check	OK	03:45	

Data Download Date/Time: 03/18/2019 17:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1048227
DRY GAS EXPIRY DATE: 11/15/2021
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: FEES

DATE/TIME OF TEST: 02/14/2019 00:56:39

Diagnostics Check	OK	00:48
Air Blank	0.000	00:49
Control Test	0.084	00:49
Air Blank	0.000	00:50
Subject Sample # 1	0.215	00:51
Breath Volume	2.410	00:51
Air Blank	0.000	00:51
Air Blank	0.000	00:53
Subject Sample # 2	0.195	00:55
Breath Volume	2.675	00:55
Air Blank	0.000	00:55
Control Test	0.084	00:56
Air Blank	0.000	00:56
Diagnostics Check	OK	00:56

Data Download Date/Time: 02/27/2019 19:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1048227
DRY GAS EXPIRY DATE: 11/15/2021
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: ORSELLI JR

DATE/TIME OF TEST: 02/24/2019 00:09:23

Diagnostics Check	OK	00:00
Air Blank	0.000	00:01
Control Test	0.083	00:01
Air Blank	0.000	00:02
Subject Sample # 1	0.133	00:03
Breath Volume	2.066	00:03
Air Blank	0.000	00:04
Air Blank	0.000	00:05
Subject Sample # 2	0.116	00:07
Breath Volume	1.167	00:07
Air Blank	0.000	00:08
Control Test	0.083	00:08
Air Blank	0.000	00:09
Diagnostics Check	OK	00:09

Data Download Date/Time: 02/27/2019 19:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1048227
DRY GAS EXPIRY DATE: 11/15/2021
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: GOMEZ GONZALEZ

DATE/TIME OF TEST: 02/24/2019 03:25:10

Diagnostics Check	OK	03:18
Air Blank	0.000	03:18
Control Test	0.083	03:18
Air Blank	0.000	03:19
Subject Sample # 1	0.171	03:20
Breath Volume	1.121	03:20
Air Blank	0.000	03:20
Air Blank	0.000	03:22
Subject Sample # 2	0.174	03:23
Breath Volume	1.957	03:23
Air Blank	0.000	03:24
Control Test	0.082	03:24
Air Blank	0.000	03:25
Diagnostics Check	OK	03:25

Data Download Date/Time: 02/27/2019 19:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WALTER,CRAIG,K

BREATH TEST OPERATOR AGENCY: Deland PD

ARRESTING OFFICER: THORNTON,K

ARRESTING OFFICER AGENCY: FFWCC (Ocala)

LAST AGENCY INSPECTION: 01/30/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 970670

DRY GAS EXPIRY DATE: 06/07/2021

VIOLATION CODE: DUI

ARREST TIME: 19:27

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:26

SUBJECT'S LAST NAME: TAUBE

DATE/TIME OF TEST: 02/01/2019 20:57:51

Diagnostics Check	OK	20:50
Air Blank	0.000	20:50
Control Test	0.080	20:50
Air Blank	0.000	20:51
Subject Sample # 1	0.226	20:52
Breath Volume	1.652	20:52
Air Blank	0.000	20:53
Air Blank	0.000	20:55
Subject Sample # 2	0.211	20:56
Breath Volume	1.281	20:56
Air Blank	0.000	20:56
Control Test	0.079	20:57
Air Blank	0.000	20:57
Diagnostics Check	OK	20:57

Data Download Date/Time: 02/28/2019 16:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALTER,CRAIG,K
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: WINLAND,BOBBY
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 01/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: PEDRAZA

DATE/TIME OF TEST: 02/03/2019 02:31:51

Diagnostics Check	OK	02:28	
Air Blank	0.000	02:29	
Control Test	0.080	02:29	
Air Blank	0.000	02:29	
Subject Sample	# 1	02:30	Subject Test Refused
Breath Volume	0.000	02:30	
Air Blank	0.000	02:30	
Control Test	0.079	02:31	
Air Blank	0.000	02:31	
Diagnostics Check	OK	02:31	

Data Download Date/Time: 02/28/2019 16:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAGLE, TONY,
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: ROYCE, JAMES
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 01/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: HART

DATE/TIME OF TEST: 02/14/2019 01:22:12

Diagnostics Check	OK	01:15
Air Blank	0.000	01:15
Control Test	0.080	01:15
Air Blank	0.000	01:16
Subject Sample # 1	0.009	01:17
Breath Volume	3.441	01:17
Air Blank	0.000	01:17
Air Blank	0.000	01:19
Subject Sample # 2	0.009	01:20
Breath Volume	4.558	01:20
Air Blank	0.000	01:21
Control Test	0.081	01:21
Air Blank	0.000	01:22
Diagnostics Check	OK	01:22

Data Download Date/Time: 02/28/2019 16:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAGLE, TONY,
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: BURGER, ALISHA
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 01/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: SHAFFER

DATE/TIME OF TEST: 02/15/2019 04:41:00

Diagnostics Check	OK	04:34
Air Blank	0.000	04:35
Control Test	0.080	04:35
Air Blank	0.000	04:36
Subject Sample # 1	0.249	04:36
Breath Volume	2.890	04:36
Air Blank	0.000	04:37
Air Blank	0.000	04:38
Subject Sample # 2	0.266	04:39
Breath Volume	3.472	04:39
Air Blank	0.000	04:40
Control Test	0.079	04:40
Air Blank	0.000	04:40
Diagnostics Check	OK	04:41

Data Download Date/Time: 02/28/2019 16:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAGLE, TONY,
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: ORITZ, ROBERTO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: AGUILAR JUAREZ

DATE/TIME OF TEST: 02/20/2019 05:34:19

Diagnostics Check	OK	05:27
Air Blank	0.000	05:27
Control Test	0.080	05:27
Air Blank	0.000	05:28
Subject Sample # 1	0.213	05:29
Breath Volume	2.394	05:29
Air Blank	0.000	05:29
Air Blank	0.000	05:31
Subject Sample # 2	0.229	05:32
Breath Volume	2.121	05:32
Air Blank	0.000	05:33
Control Test	0.080	05:33
Air Blank	0.000	05:34
Diagnostics Check	OK	05:34

Data Download Date/Time: 02/28/2019 16:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WALTER,CRAIG,K

BREATH TEST OPERATOR AGENCY: Deland PD

ARRESTING OFFICER: JAMES,ROYCE

ARRESTING OFFICER AGENCY: Volusia County SO

LAST AGENCY INSPECTION: 01/30/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 970670

DRY GAS EXPIRY DATE: 06/07/2021

VIOLATION CODE: DUI

ARREST TIME: 19:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: ALQUISIRA

DATE/TIME OF TEST: 02/20/2019 20:43:33

Diagnostics Check	OK	20:37
Air Blank	0.000	20:37
Control Test	0.080	20:37
Air Blank	0.000	20:38
Subject Sample # 1	0.324	20:38
Breath Volume	3.144	20:38
Air Blank	0.000	20:39
Air Blank	0.000	20:41
Subject Sample # 2	0.322	20:41
Breath Volume	3.320	20:41
Air Blank	0.000	20:42
Control Test	0.078	20:42
Air Blank	0.000	20:43
Diagnostics Check	OK	20:43

Data Download Date/Time: 02/28/2019 16:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILLAN, THOMAS, C
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: JONES, CALEB
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 01/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 16:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:57

SUBJECT'S LAST NAME: BROWNING

DATE/TIME OF TEST: 02/21/2019 16:34:44

Diagnostics Check	OK	16:27
Air Blank	0.000	16:28
Control Test	0.080	16:28
Air Blank	0.000	16:29
Subject Sample # 1	0.258	16:29
Breath Volume	1.238	16:29
Air Blank	0.000	16:30
Air Blank	0.000	16:32
Subject Sample # 2	0.256	16:33
Breath Volume	1.277	16:33
Air Blank	0.000	16:33
Control Test	0.079	16:34
Air Blank	0.000	16:34
Diagnostics Check	OK	16:34

Data Download Date/Time: 02/28/2019 16:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAGLE, TONY,
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: CODY, RICHARD
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 01/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 18:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: DEARBECK

DATE/TIME OF TEST: 02/22/2019 19:19:25

Diagnostics Check	OK	19:16	
Air Blank	0.000	19:16	
Control Test	0.079	19:17	
Air Blank	0.000	19:17	
Subject Sample	# 1	19:18	Subject Test Refused
Breath Volume	0.000	19:18	
Air Blank	0.000	19:18	
Control Test	0.079	19:18	
Air Blank	0.000	19:19	
Diagnostics Check	OK	19:19	

Data Download Date/Time: 02/28/2019 16:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAGLE, TONY,
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: TAGLE, TONY
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 01/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: FLORES OLASCUAGA

DATE/TIME OF TEST: 02/23/2019 23:42:26

Diagnostics Check	OK	23:36
Air Blank	0.000	23:36
Control Test	0.079	23:36
Air Blank	0.000	23:37
Subject Sample # 1	0.182	23:37
Breath Volume	2.718	23:37
Air Blank	0.000	23:38
Air Blank	0.000	23:40
Subject Sample # 2	0.180	23:40
Breath Volume	2.992	23:40
Air Blank	0.000	23:41
Control Test	0.080	23:41
Air Blank	0.000	23:42
Diagnostics Check	OK	23:42

Data Download Date/Time: 02/28/2019 16:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAGLE, TONY,
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: STACY, CODY
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 01/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 04:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: NORIEGA

DATE/TIME OF TEST: 02/24/2019 05:43:51

Diagnostics Check	OK	05:37
Air Blank	0.000	05:38
Control Test	0.080	05:38
Air Blank	0.000	05:39
Subject Sample # 1	0.129	05:39
Breath Volume	2.894	05:39
Air Blank	0.000	05:40
Air Blank	0.000	05:41
Subject Sample # 2	0.122	05:42
Breath Volume	3.117	05:42
Air Blank	0.000	05:42
Control Test	0.079	05:43
Air Blank	0.000	05:43
Diagnostics Check	OK	05:43

Data Download Date/Time: 02/28/2019 16:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAGLE, TONY,
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: BURGER, ALISHA
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 01/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 02/25/2019 00:47:56

Diagnostics Check	OK	00:45	
Air Blank	0.000	00:45	
Control Test	0.080	00:45	
Air Blank	0.000	00:46	
Subject Sample	# 1	00:46	Subject Test Refused
Breath Volume	0.000	00:46	
Air Blank	0.000	00:47	
Control Test	0.080	00:47	
Air Blank	0.000	00:47	
Diagnostics Check	OK	00:47	

Data Download Date/Time: 02/28/2019 16:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAGLE, TONY,
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: WELCH, KYLE
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 01/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: BORBON

DATE/TIME OF TEST: 02/25/2019 03:08:27

Diagnostics Check	OK	03:01
Air Blank	0.000	03:01
Control Test	0.081	03:01
Air Blank	0.000	03:02
Subject Sample # 1	0.033	03:03
Breath Volume	3.839	03:03
Air Blank	0.000	03:04
Air Blank	0.000	03:06
Subject Sample # 2	0.029	03:07
Breath Volume	2.011	03:07
Air Blank	0.000	03:07
Control Test	0.080	03:07
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

Data Download Date/Time: 02/28/2019 16:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESTELUS,JUNIOR,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: ZIGLER,RICKIE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 437738
DRY GAS EXPIRY DATE: 06/08/2019
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: GAYTAN-MEDINA

DATE/TIME OF TEST: 02/24/2019 03:33:13

Diagnostics Check	OK	03:25
Air Blank	0.000	03:26
Control Test	0.081	03:26
Air Blank	0.000	03:27
Subject Sample # 1	0.141	03:28
Breath Volume	2.132	03:28
Air Blank	0.000	03:29
Air Blank	0.000	03:31
Subject Sample # 2	0.144	03:31
Breath Volume	2.183	03:31
Air Blank	0.000	03:32
Control Test	0.080	03:32
Air Blank	0.000	03:33
Diagnostics Check	OK	03:33

Data Download Date/Time: 02/25/2019 18:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001122

INSTRUMENT REGISTERED TO: FHP MIAMI

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FRANCESCHI,VANESSA,R

BREATH TEST OPERATOR AGENCY: FHP Troop K

ARRESTING OFFICER: FRANCESCHI,VANESSA

ARRESTING OFFICER AGENCY: FHP Troop K

LAST AGENCY INSPECTION: 01/02/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 796576

DRY GAS EXPIRY DATE: 06/14/2020

VIOLATION CODE: DUI

ARREST TIME: 13:43

GENDER: Female

OBSERVATION PERIOD BEGAN: 14:08

SUBJECT'S LAST NAME: GALVIS

DATE/TIME OF TEST: 02/13/2019 14:31:40

Diagnostics Check	OK	14:28	
Air Blank	0.000	14:29	
Control Test	0.078	14:29	
Air Blank	0.000	14:30	
Subject Sample	# 1	14:30	Subject Test Refused
Breath Volume	0.000	14:30	
Air Blank	0.000	14:30	
Control Test	0.078	14:31	
Air Blank	0.000	14:31	
Diagnostics Check	OK	14:31	

Data Download Date/Time: 02/27/2019 09:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001123

INSTRUMENT REGISTERED TO: FHP TROOP D ORLANDO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRACH,STEFAN,G
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: DRACH,STEFAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: NARCISSE

DATE/TIME OF TEST: 02/07/2019 03:50:40

Diagnostics Check	OK	03:44	
Air Blank	0.000	03:44	
Control Test	0.084	03:44	
Air Blank	0.000	03:45	
Subject Sample # 1	0.074	03:46	
Breath Volume	2.191	03:46	
Air Blank	0.000	03:46	
Air Blank	0.000	03:48	
Subject Sample # 2	0.081	03:49	
Breath Volume	3.871	03:49	
Air Blank	0.000	03:49	
Control Test	0.000	03:50	Control Outside Tolerance
Air Blank	0.000	03:50	

Data Download Date/Time: 02/18/2019 19:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001123

INSTRUMENT REGISTERED TO: FHP TROOP D ORLANDO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRACH,STEFAN,G
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: DRACH,STEFAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: NARCISSE

DATE/TIME OF TEST: 02/07/2019 04:34:38

Diagnostics Check	OK	04:26
Air Blank	0.000	04:26
Control Test	0.079	04:26
Air Blank	0.000	04:27
Subject Sample # 1	0.071	04:30
Breath Volume	3.457	04:30
Air Blank	0.000	04:30
Air Blank	0.000	04:32
Subject Sample # 2	0.069	04:33
Breath Volume	3.527	04:33
Air Blank	0.000	04:33
Control Test	0.079	04:34
Air Blank	0.000	04:34
Diagnostics Check	OK	04:34

Data Download Date/Time: 02/18/2019 19:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001123

INSTRUMENT REGISTERED TO: FHP TROOP D ORLANDO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRACH,STEFAN,G
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: DRACH,STEFAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: ABDELHACK

DATE/TIME OF TEST: 02/16/2019 00:57:09

Diagnostics Check	OK	00:47
Air Blank	0.000	00:47
Control Test	0.079	00:48
Air Blank	0.000	00:48
Subject Sample # 1	0.155	00:49
Breath Volume	1.128	00:49
Air Blank	0.000	00:50
Air Blank	0.000	00:52
Subject Sample # 2	0.166	00:55
Breath Volume	2.750	00:55
Air Blank	0.000	00:56
Control Test	0.080	00:56
Air Blank	0.000	00:57
Diagnostics Check	OK	00:57

Data Download Date/Time: 02/18/2019 19:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BING, TRAE, J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: COGGINS, JACHIN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 02/02/2019 03:37:14

Diagnostics Check	OK	03:24	
Air Blank	0.000	03:25	
Control Test	0.079	03:25	
Air Blank	0.000	03:26	
Subject Sample	# 1	03:29	No Sample Provided
Breath Volume	0.000	03:29	
Air Blank	0.000	03:29	
Air Blank	0.000	03:31	
Subject Sample	# 2	0.110	03:32
Breath Volume	2.199	03:32	
Air Blank	0.000	03:33	
Air Blank	0.000	03:35	
Subject Sample	# 3	0.112	03:35
Breath Volume	1.621	03:35	
Air Blank	0.000	03:36	
Control Test	0.078	03:36	
Air Blank	0.000	03:37	
Diagnostics Check	OK	03:37	

Data Download Date/Time: 02/27/2019 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES, STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SYKES, STEVEN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: WITHEY

DATE/TIME OF TEST: 02/03/2019 00:03:36

Diagnostics Check	OK	23:52	
Air Blank	0.000	23:52	
Control Test	0.080	23:53	
Air Blank	0.000	23:53	
Subject Sample	# 2	23:56	No Sample Provided
Breath Volume	0.000	23:56	
Air Blank	0.000	23:57	
Air Blank	0.000	23:59	
Subject Sample	# 1	00:02	No Sample Provided
Breath Volume	0.000	00:02	
Air Blank	0.000	00:02	
Control Test	0.079	00:03	
Air Blank	0.000	00:03	
Diagnostics Check	OK	00:03	

Data Download Date/Time: 02/27/2019 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHRISTELLO, CLIFFORD, A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HAMILTON, ROBERT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 02/03/2019 01:00:41

Diagnostics Check	OK	00:53
Air Blank	0.000	00:53
Control Test	0.081	00:54
Air Blank	0.000	00:54
Subject Sample # 1	0.175	00:55
Breath Volume	1.738	00:55
Air Blank	0.000	00:56
Air Blank	0.000	00:58
Subject Sample # 2	0.169	00:59
Breath Volume	1.390	00:59
Air Blank	0.000	00:59
Control Test	0.080	01:00
Air Blank	0.000	01:00
Diagnostics Check	OK	01:00

Data Download Date/Time: 02/27/2019 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILLIPS,JORDAN,W
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: FRANCE,PAUL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:55

SUBJECT'S LAST NAME: PEREZ-PEREZ

DATE/TIME OF TEST: 02/03/2019 06:25:45

Diagnostics Check	OK	06:18
Air Blank	0.000	06:19
Control Test	0.079	06:19
Air Blank	0.000	06:20
Subject Sample # 1	0.174	06:21
Breath Volume	2.781	06:21
Air Blank	0.000	06:21
Air Blank	0.000	06:23
Subject Sample # 2	0.162	06:24
Breath Volume	2.273	06:24
Air Blank	0.000	06:24
Control Test	0.079	06:25
Air Blank	0.000	06:25
Diagnostics Check	OK	06:25

Data Download Date/Time: 02/27/2019 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOTO,MAVERICK,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SOTO,MAVERICK
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: BRYAN

DATE/TIME OF TEST: 02/03/2019 22:59:31

Diagnostics Check	OK	22:53
Air Blank	0.000	22:54
Control Test	0.080	22:54
Air Blank	0.000	22:54
Subject Sample # 1	0.026	22:55
Breath Volume	2.566	22:55
Air Blank	0.000	22:55
Air Blank	0.000	22:57
Subject Sample # 2	0.024	22:58
Breath Volume	3.304	22:58
Air Blank	0.000	22:58
Control Test	0.079	22:58
Air Blank	0.000	22:59
Diagnostics Check	OK	22:59

Data Download Date/Time: 02/27/2019 11:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JEUDY,RINALDO,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HIGGINBOTHAM,JACK
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:08

SUBJECT'S LAST NAME: OBRIEN

DATE/TIME OF TEST: 02/03/2019 23:25:04

Diagnostics Check	OK	23:22	
Air Blank	0.000	23:22	
Control Test	0.079	23:22	
Air Blank	0.000	23:23	
Subject Sample	# 1	23:23	Subject Test Refused
Breath Volume	0.000	23:23	
Air Blank	0.000	23:24	
Control Test	0.080	23:24	
Air Blank	0.000	23:24	
Diagnostics Check	OK	23:25	

Data Download Date/Time: 02/27/2019 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES,STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SYKES,STEVEN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: GUY

DATE/TIME OF TEST: 02/04/2019 00:59:05

Diagnostics Check	OK	00:52
Air Blank	0.000	00:52
Control Test	0.080	00:53
Air Blank	0.000	00:53
Subject Sample # 1	0.155	00:54
Breath Volume	1.699	00:54
Air Blank	0.000	00:55
Air Blank	0.000	00:57
Subject Sample # 2	0.153	00:57
Breath Volume	1.351	00:57
Air Blank	0.000	00:58
Control Test	0.079	00:58
Air Blank	0.000	00:59
Diagnostics Check	OK	00:59

Data Download Date/Time: 02/27/2019 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES,STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HAMILTON,ROBERT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: ZEFERINO

DATE/TIME OF TEST: 02/04/2019 01:58:19

Diagnostics Check	OK	01:52
Air Blank	0.000	01:52
Control Test	0.079	01:52
Air Blank	0.000	01:53
Subject Sample # 1	0.115	01:53
Breath Volume	1.250	01:53
Air Blank	0.000	01:54
Air Blank	0.000	01:56
Subject Sample # 2	0.119	01:56
Breath Volume	2.261	01:56
Air Blank	0.000	01:57
Control Test	0.080	01:57
Air Blank	0.000	01:58
Diagnostics Check	OK	01:58

Data Download Date/Time: 02/27/2019 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCH,JOHN,M
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: CHURCH,JOHN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 10:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:01

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 02/04/2019 11:28:00

Diagnostics Check	OK	11:21
Air Blank	0.000	11:22
Control Test	0.080	11:22
Air Blank	0.000	11:23
Subject Sample # 1	0.000	11:23
Breath Volume	2.250	11:23
Air Blank	0.000	11:24
Air Blank	0.000	11:26
Subject Sample # 2	0.000	11:26
Breath Volume	1.816	11:26
Air Blank	0.000	11:27
Control Test	0.080	11:27
Air Blank	0.000	11:27
Diagnostics Check	OK	11:28

Data Download Date/Time: 02/27/2019 11:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEWETT,MICHAEL,W
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HEWETT,MICHAEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 15:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:05

SUBJECT'S LAST NAME: CHILDERS

DATE/TIME OF TEST: 02/07/2019 16:35:40

Diagnostics Check	OK	16:29
Air Blank	0.000	16:30
Control Test	0.079	16:30
Air Blank	0.000	16:30
Subject Sample # 1	0.266	16:31
Breath Volume	2.394	16:31
Air Blank	0.000	16:31
Air Blank	0.000	16:33
Subject Sample # 2	0.257	16:34
Breath Volume	2.527	16:34
Air Blank	0.000	16:34
Control Test	0.079	16:35
Air Blank	0.000	16:35
Diagnostics Check	OK	16:35

Data Download Date/Time: 02/27/2019 11:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JEUDY,RINALDO,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HAMILTON,ROBERT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 02/08/2019 04:46:13

Diagnostics Check	OK	04:43	
Air Blank	0.000	04:43	
Control Test	0.079	04:44	
Air Blank	0.000	04:44	
Subject Sample	# 1	04:44	Subject Test Refused
Breath Volume	0.000	04:44	
Air Blank	0.000	04:45	
Control Test	0.080	04:45	
Air Blank	0.000	04:46	
Diagnostics Check	OK	04:46	

Data Download Date/Time: 02/27/2019 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,ROBERT,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: CARTER,JACOB
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 01/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: BUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: BRIGGS

DATE/TIME OF TEST: 02/09/2019 02:01:04

Diagnostics Check	OK	01:58	
Air Blank	0.000	01:58	
Control Test	0.080	01:59	
Air Blank	0.000	01:59	
Subject Sample	# 1	01:59	Subject Test Refused
Breath Volume	0.000	01:59	
Air Blank	0.000	02:00	
Control Test	0.080	02:00	
Air Blank	0.000	02:00	
Diagnostics Check	OK	02:01	

Data Download Date/Time: 02/27/2019 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMMONDS,TEAQULIA,L
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:22

SUBJECT'S LAST NAME: HERNANDEZ BRUNO

DATE/TIME OF TEST: 02/09/2019 06:47:58

Diagnostics Check	OK	06:43	
Air Blank	0.000	06:44	
Control Test	0.080	06:44	
Air Blank	0.000	06:44	
Subject Sample	# 1	06:46	Subject Test Refused
Breath Volume	0.257	06:46	
Air Blank	0.000	06:47	
Control Test	0.080	06:47	
Air Blank	0.000	06:47	
Diagnostics Check	OK	06:47	

Data Download Date/Time: 02/27/2019 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COKE,CLEVELAND,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: CARTER,JACOB
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 01/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: HAYES

DATE/TIME OF TEST: 02/10/2019 01:57:38

Diagnostics Check	OK	01:51
Air Blank	0.000	01:52
Control Test	0.080	01:52
Air Blank	0.000	01:52
Subject Sample # 1	0.021	01:53
Breath Volume	2.847	01:53
Air Blank	0.000	01:53
Air Blank	0.000	01:55
Subject Sample # 2	0.019	01:56
Breath Volume	2.683	01:56
Air Blank	0.000	01:56
Control Test	0.079	01:57
Air Blank	0.000	01:57
Diagnostics Check	OK	01:57

Data Download Date/Time: 02/27/2019 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHRISTELLO,CLIFFORD,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: CHIRSTELLO,CLIFFORD
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:04

SUBJECT'S LAST NAME: HUDDLESON

DATE/TIME OF TEST: 02/11/2019 21:21:18

Diagnostics Check	OK	21:09	
Air Blank	0.000	21:10	
Control Test	0.080	21:10	
Air Blank	0.000	21:11	
Subject Sample	# 1	21:14	No Sample Provided
Breath Volume	0.000	21:14	
Air Blank	0.000	21:14	
Air Blank	0.000	21:16	
Subject Sample	# 2	21:19	No Sample Provided
Breath Volume	0.000	21:19	
Air Blank	0.000	21:20	
Control Test	0.081	21:20	
Air Blank	0.000	21:21	
Diagnostics Check	OK	21:21	

Data Download Date/Time: 02/27/2019 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHRISTELLO,CLIFFORD,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: CHRISTELLO,CLIFFORD
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: DOBLADO

DATE/TIME OF TEST: 02/12/2019 03:02:36

Diagnostics Check	OK	02:56
Air Blank	0.000	02:56
Control Test	0.080	02:57
Air Blank	0.000	02:57
Subject Sample # 1	0.223	02:58
Breath Volume	2.019	02:58
Air Blank	0.000	02:58
Air Blank	0.000	03:00
Subject Sample # 2	0.226	03:01
Breath Volume	1.972	03:01
Air Blank	0.000	03:01
Control Test	0.080	03:02
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

Data Download Date/Time: 02/27/2019 11:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAWKINS,KIMBERLY,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: MATHIS,JORDAN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 07:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: CREWS

DATE/TIME OF TEST: 02/15/2019 09:25:01

Diagnostics Check	OK	09:21	
Air Blank	0.000	09:22	
Control Test	0.080	09:22	
Air Blank	0.000	09:23	
Subject Sample	# 1	09:23	Subject Test Refused
Breath Volume	0.000	09:23	
Air Blank	0.000	09:24	
Control Test	0.080	09:24	
Air Blank	0.000	09:24	
Diagnostics Check	OK	09:25	

Data Download Date/Time: 02/27/2019 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JEUDY,RINALDO,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HIGGINBOTHAM,JACK
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: BARNES

DATE/TIME OF TEST: 02/16/2019 02:24:58

Diagnostics Check	OK	02:12	
Air Blank	0.000	02:12	
Control Test	0.079	02:12	
Air Blank	0.000	02:13	
Subject Sample	# 1 0.135	02:16	Volume Not Met
Breath Volume	1.082	02:16	
Air Blank	0.000	02:17	
Air Blank	0.000	02:18	
Subject Sample	# 2 0.133	02:20	
Breath Volume	1.722	02:20	
Air Blank	0.000	02:20	
Air Blank	0.000	02:22	
Subject Sample	# 3 0.141	02:23	
Breath Volume	1.152	02:23	
Air Blank	0.000	02:24	
Control Test	0.078	02:24	
Air Blank	0.000	02:24	
Diagnostics Check	OK	02:24	

Data Download Date/Time: 02/27/2019 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JEUDY,RINALDO,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HAMILTON,ROBERT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: CORTES MONTERO

DATE/TIME OF TEST: 02/16/2019 03:25:50

Diagnostics Check	OK	03:19
Air Blank	0.000	03:19
Control Test	0.079	03:20
Air Blank	0.000	03:20
Subject Sample # 1	0.191	03:21
Breath Volume	2.222	03:21
Air Blank	0.000	03:21
Air Blank	0.000	03:23
Subject Sample # 2	0.183	03:24
Breath Volume	1.972	03:24
Air Blank	0.000	03:24
Control Test	0.079	03:25
Air Blank	0.000	03:25
Diagnostics Check	OK	03:25

Data Download Date/Time: 02/27/2019 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHRISTELLO, CLIFFORD, A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HAMILTON, ROBERT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: VANN

DATE/TIME OF TEST: 02/17/2019 00:53:54

Diagnostics Check	OK	00:42	
Air Blank	0.000	00:43	
Control Test	0.080	00:43	
Air Blank	0.000	00:43	
Subject Sample	# 1	00:46	No Sample Provided
Breath Volume	0.000	00:46	
Air Blank	0.000	00:47	
Air Blank	0.000	00:49	
Subject Sample	# 2	00:52	No Sample Provided
Breath Volume	0.000	00:52	
Air Blank	0.000	00:53	
Control Test	0.080	00:53	
Air Blank	0.000	00:53	
Diagnostics Check	OK	00:53	

Data Download Date/Time: 02/27/2019 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHRISTELLO,CLIFFORD,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: CHRISTELLO,CLIFFORD
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: DYKES

DATE/TIME OF TEST: 02/17/2019 01:17:57

Diagnostics Check	OK	01:06	
Air Blank	0.000	01:07	
Control Test	0.080	01:07	
Air Blank	0.000	01:07	
Subject Sample	# 1	01:11	No Sample Provided
Breath Volume	0.000	01:11	
Air Blank	0.000	01:11	
Air Blank	0.000	01:13	
Subject Sample	# 2	01:16	No Sample Provided
Breath Volume	0.000	01:16	
Air Blank	0.000	01:17	
Control Test	0.081	01:17	
Air Blank	0.000	01:17	
Diagnostics Check	OK	01:17	

Data Download Date/Time: 02/27/2019 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JEUDY,RINALDO,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HAMILTON,ROBERT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: BELTZ

DATE/TIME OF TEST: 02/18/2019 02:04:06

Diagnostics Check	OK	01:55
Air Blank	0.000	01:55
Control Test	0.078	01:55
Air Blank	0.000	01:56
Subject Sample # 1	0.166	01:59
Breath Volume	1.589	01:59
Air Blank	0.000	01:59
Air Blank	0.000	02:01
Subject Sample # 2	0.169	02:02
Breath Volume	1.148	02:02
Air Blank	0.000	02:03
Control Test	0.078	02:03
Air Blank	0.000	02:04
Diagnostics Check	OK	02:04

Data Download Date/Time: 02/27/2019 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BING, TRAE, J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HIGGINBOTHAM, JACK
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:43

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 02/20/2019 21:37:18

Diagnostics Check	OK	21:25	
Air Blank	0.000	21:25	
Control Test	0.079	21:25	
Air Blank	0.000	21:26	
Subject Sample	# 1 0.038	21:29	Volume Not Met
Breath Volume	0.191	21:29	
Air Blank	0.000	21:30	
Air Blank	0.000	21:32	
Subject Sample	# 2 0.034	21:32	
Breath Volume	1.437	21:32	
Air Blank	0.000	21:33	
Air Blank	0.000	21:35	
Subject Sample	# 3 0.035	21:35	
Breath Volume	1.789	21:35	
Air Blank	0.000	21:36	
Control Test	0.080	21:36	
Air Blank	0.000	21:37	
Diagnostics Check	OK	21:37	

Data Download Date/Time: 02/27/2019 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JEUDY,RINALDO,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: DAVIS,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: SPIKER

DATE/TIME OF TEST: 02/21/2019 04:18:07

Diagnostics Check	OK	04:11
Air Blank	0.000	04:12
Control Test	0.080	04:12
Air Blank	0.000	04:12
Subject Sample # 1	0.203	04:13
Breath Volume	2.433	04:13
Air Blank	0.000	04:14
Air Blank	0.000	04:16
Subject Sample # 2	0.208	04:16
Breath Volume	2.902	04:16
Air Blank	0.000	04:17
Control Test	0.079	04:17
Air Blank	0.000	04:18
Diagnostics Check	OK	04:18

Data Download Date/Time: 02/27/2019 11:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JEUDY,RINALDO,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HIGGINBOTHAM,JACK
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: LANIER

DATE/TIME OF TEST: 02/21/2019 22:43:49

Diagnostics Check	OK	22:41	
Air Blank	0.000	22:41	
Control Test	0.080	22:41	
Air Blank	0.000	22:42	
Subject Sample	# 1	22:42	Subject Test Refused
Breath Volume	0.000	22:42	
Air Blank	0.000	22:42	
Control Test	0.080	22:43	
Air Blank	0.000	22:43	
Diagnostics Check	OK	22:43	

Data Download Date/Time: 02/27/2019 11:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELTON,XAVIER,V
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 02/22/2019 22:16:34

Diagnostics Check	OK	22:10
Air Blank	0.000	22:10
Control Test	0.080	22:10
Air Blank	0.000	22:11
Subject Sample # 1	0.201	22:12
Breath Volume	2.593	22:12
Air Blank	0.000	22:12
Air Blank	0.000	22:14
Subject Sample # 2	0.193	22:15
Breath Volume	2.945	22:15
Air Blank	0.000	22:15
Control Test	0.078	22:15
Air Blank	0.000	22:16
Diagnostics Check	OK	22:16

Data Download Date/Time: 02/27/2019 11:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELTON,XAVIER,V
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ROBERTS,DALE
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: DRAWHORN

DATE/TIME OF TEST: 02/23/2019 23:34:20

Diagnostics Check	OK	23:31	
Air Blank	0.000	23:31	
Control Test	0.080	23:32	
Air Blank	0.000	23:32	
Subject Sample	# 1	23:32	Subject Test Refused
Breath Volume	0.000	23:32	
Air Blank	0.000	23:33	
Control Test	0.080	23:33	
Air Blank	0.000	23:34	
Diagnostics Check	OK	23:34	

Data Download Date/Time: 02/27/2019 11:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELTON,XAVIER,V
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,A
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 02/24/2019 04:39:55

Diagnostics Check	OK	04:33
Air Blank	0.000	04:33
Control Test	0.080	04:34
Air Blank	0.000	04:34
Subject Sample # 1	0.192	04:34
Breath Volume	1.609	04:34
Air Blank	0.000	04:35
Air Blank	0.000	04:37
Subject Sample # 2	0.183	04:38
Breath Volume	1.839	04:38
Air Blank	0.000	04:39
Control Test	0.081	04:39
Air Blank	0.000	04:39
Diagnostics Check	OK	04:39

Data Download Date/Time: 02/27/2019 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHRISTELLO,CLIFFORD,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: CHRISTELLO,CLIFFORD
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 15:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:24

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 02/25/2019 16:51:21

Diagnostics Check	OK	16:44
Air Blank	0.000	16:45
Control Test	0.080	16:45
Air Blank	0.000	16:46
Subject Sample # 1	0.235	16:47
Breath Volume	1.246	16:47
Air Blank	0.000	16:47
Air Blank	0.000	16:49
Subject Sample # 2	0.235	16:49
Breath Volume	1.714	16:49
Air Blank	0.000	16:50
Control Test	0.080	16:50
Air Blank	0.000	16:51
Diagnostics Check	OK	16:51

Data Download Date/Time: 02/27/2019 11:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BING, TRAE, J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HAMILTON, ROBERT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: HAWKINS

DATE/TIME OF TEST: 02/25/2019 23:17:57

Diagnostics Check	OK	23:15	
Air Blank	0.000	23:15	
Control Test	0.080	23:15	
Air Blank	0.000	23:16	
Subject Sample	# 1	23:16	Subject Test Refused
Breath Volume	0.000	23:16	
Air Blank	0.000	23:17	
Control Test	0.080	23:17	
Air Blank	0.000	23:17	
Diagnostics Check	OK	23:17	

Data Download Date/Time: 02/27/2019 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BING, TRAE, J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HAMILTON, ROBERT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: EISSING

DATE/TIME OF TEST: 02/27/2019 03:03:05

Diagnostics Check	OK	02:56
Air Blank	0.000	02:57
Control Test	0.081	02:57
Air Blank	0.000	02:57
Subject Sample # 1	0.167	02:58
Breath Volume	1.531	02:58
Air Blank	0.000	02:59
Air Blank	0.000	03:01
Subject Sample # 2	0.177	03:01
Breath Volume	1.429	03:01
Air Blank	0.000	03:02
Control Test	0.079	03:02
Air Blank	0.000	03:03
Diagnostics Check	OK	03:03

Data Download Date/Time: 02/27/2019 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,ROBERT,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 02/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:08

SUBJECT'S LAST NAME: CHATTELLE

DATE/TIME OF TEST: 02/28/2019 21:37:13

Diagnostics Check	OK	21:29
Air Blank	0.000	21:30
Control Test	0.079	21:30
Air Blank	0.000	21:30
Subject Sample # 1	0.315	21:31
Breath Volume	2.015	21:31
Air Blank	0.000	21:32
Air Blank	0.000	21:33
Subject Sample # 2	0.316	21:35
Breath Volume	2.562	21:35
Air Blank	0.000	21:36
Control Test	0.081	21:36
Air Blank	0.000	21:37
Diagnostics Check	OK	21:37

Data Download Date/Time: 03/23/2019 10:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRYAN,STEVEN,A
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: KING,BRANDON
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 02/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: ROBLES-MULL

DATE/TIME OF TEST: 02/23/2019 03:30:14

Diagnostics Check	OK	03:16	
Air Blank	0.000	03:17	
Control Test	0.080	03:17	
Air Blank	0.000	03:17	
Subject Sample	# 1	03:21	No Sample Provided
Breath Volume	0.000	03:21	
Air Blank	0.000	03:21	
Air Blank	0.000	03:23	
Subject Sample	# 2	0.131	03:25
Breath Volume	1.335	03:25	
Air Blank	0.000	03:26	
Air Blank	0.000	03:27	
Subject Sample	# 3	0.140	03:28
Breath Volume	2.765	03:28	
Air Blank	0.000	03:29	
Control Test	0.079	03:29	
Air Blank	0.000	03:30	
Diagnostics Check	OK	03:30	

Data Download Date/Time: 03/26/2019 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRYAN,STEVEN,A
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: KING,BRANDON
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 02/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: TREMBLAY

DATE/TIME OF TEST: 02/25/2019 01:10:34

Diagnostics Check	OK	01:04
Air Blank	0.000	01:04
Control Test	0.080	01:04
Air Blank	0.000	01:05
Subject Sample # 1	0.289	01:06
Breath Volume	1.898	01:06
Air Blank	0.000	01:06
Air Blank	0.000	01:08
Subject Sample # 2	0.278	01:09
Breath Volume	2.375	01:09
Air Blank	0.000	01:09
Control Test	0.080	01:09
Air Blank	0.000	01:10
Diagnostics Check	OK	01:10

Data Download Date/Time: 03/26/2019 11:32