Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORGAN,D,
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: MORGAN,D
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: SMITHERMAN

DATE/TIME OF TEST: 04/07/2019 18:28:09

Diagnostics Check		OK	18:21
Air Blank		0.000	18:21
Control Test		0.079	18:22
Air Blank		0.000	18:22
Subject Sample	# 1	0.182	18:23
Breath Volume		2.148	18:23
Air Blank		0.000	18:24
Air Blank		0.000	18:25
Subject Sample	# 2	0.179	18:26
Breath Volume		2.281	18:26
Air Blank		0.000	18:27
Control Test		0.078	18:27
Air Blank		0.000	18:28
Diagnostics Check		OK	18:28

Data Download Date/Time: 04/28/2019 15:29

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORGAN,D,
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: MORGAN,D
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 13:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:05

SUBJECT'S LAST NAME: KARLOVA

DATE/TIME OF TEST: 04/12/2019 13:37:28

Diagnostics Check		OK	13:31
Air Blank		0.000	13:31
Control Test		0.079	13:31
Air Blank		0.000	13:32
Subject Sample	# 1	0.216	13:32
Breath Volume		2.183	13:32
Air Blank		0.000	13:33
Air Blank		0.000	13:35
Subject Sample	# 2	0.210	13:35
Breath Volume		2.308	13:35
Air Blank		0.000	13:36
Control Test		0.078	13:36
Air Blank		0.000	13:37
Diagnostics Check		OK	13:37

Data Download Date/Time: 04/28/2019 15:29

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001737

INSTRUMENT REGISTERED TO: SOUTH MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANDA,CHELSY,
BREATH TEST OPERATOR AGENCY: SOUTH MIAMI PD
ARRESTING OFFICER: LANDA,CHELSY
ARRESTING OFFICER AGENCY: SOUTH MIAMI PD
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: LIZCANO

DATE/TIME OF TEST: 04/13/2019 04:16:02

Diagnostics Check		OK	04:12	
Air Blank		0.000	04:13	
Control Test		0.079	04:13	
Air Blank		0.000	04:13	
Subject Sample	# 1		04:14	Subject Test Refused
Breath Volume		0.000	04:14	
Air Blank		0.000	04:15	
Control Test		0.079	04:15	
Air Blank		0.000	04:15	
Diagnostics Check		OK	04:16	

Data Download Date/Time: 04/26/2019 12:49

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001737

INSTRUMENT REGISTERED TO: SOUTH MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANDA,CHELSY,
BREATH TEST OPERATOR AGENCY: SOUTH MIAMI PD
ARRESTING OFFICER: MARTINEZ
HERRERA,ORESTES
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Female

OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 04/18/2019 23:34:48

Diagnostics Check		OK	23:26	
Air Blank		0.000	23:26	
Control Test		0.079	23:27	
Air Blank		0.000	23:27	
Subject Sample	# 1		23:30	No Sample Provided
Breath Volume		0.000	23:30	
Air Blank		0.000	23:31	
Air Blank		0.000	23:33	
Subject Sample	# 2		23:33	Subject Test Refused
Breath Volume		0.000	23:33	
Air Blank		0.000	23:33	
Control Test		0.078	23:34	
Air Blank		0.000	23:34	
Diagnostics Check		OK	23:34	

Data Download Date/Time: 04/26/2019 12:49

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001737

INSTRUMENT REGISTERED TO: SOUTH MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANDA,CHELSY,
BREATH TEST OPERATOR AGENCY: SOUTH MIAMI PD
ARRESTING OFFICER: LANDA,CHELSY
ARRESTING OFFICER AGENCY: SOUTH MIAMI PD
LAST AGENCY INSPECTION: 04/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: CASTANO

DATE/TIME OF TEST: 04/27/2019 03:59:18

Diagnostics Check		OK	03:53	
Air Blank		0.000	03:53	
Control Test		0.078	03:54	
Air Blank		0.000	03:54	
Subject Sample	# 1	0.169	03:57	
Breath Volume		2.992	03:57	
Air Blank			03:58	Purge Fail
Air Blank		0.000	03:59	

Data Download Date/Time: 05/30/2019 09:57

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001737

INSTRUMENT REGISTERED TO: SOUTH MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANDA,CHELSY,
BREATH TEST OPERATOR AGENCY: SOUTH MIAMI PD
ARRESTING OFFICER: LANDA,CHELSY
ARRESTING OFFICER AGENCY: SOUTH MIAMI PD
LAST AGENCY INSPECTION: 04/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: CASTANO

DATE/TIME OF TEST: 04/27/2019 04:08:12

Diagnostics Check		OK	04:01
Air Blank		0.000	04:01
Control Test		0.077	04:02
Air Blank		0.000	04:02
Subject Sample	# 1	0.173	04:03
Breath Volume		3.441	04:03
Air Blank		0.000	04:04
Air Blank		0.000	04:05
Subject Sample	# 2	0.177	04:06
Breath Volume		2.167	04:06
Air Blank		0.000	04:07
Control Test		0.077	04:07
Air Blank		0.000	04:08
Diagnostics Check		OK	04:08

Data Download Date/Time: 05/30/2019 09:58

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: CLAVETTE

DATE/TIME OF TEST: 04/02/2019 03:14:00

Diagnostics Check		OK	03:07
Air Blank		0.000	03:07
Control Test		0.078	03:07
Air Blank		0.000	03:08
Subject Sample	# 1	0.160	03:09
Breath Volume		2.996	03:09
Air Blank		0.000	03:09
Air Blank		0.000	03:11
Subject Sample	# 2	0.158	03:12
Breath Volume		2.632	03:12
Air Blank		0.000	03:13
Control Test		0.077	03:13
Air Blank		0.000	03:13
Diagnostics Check		OK	03:14

Data Download Date/Time: 04/23/2019 16:18

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: BECERRA

DATE/TIME OF TEST: 04/03/2019 00:01:20

Diagnostics Check		OK	23:58	
Air Blank		0.000	23:58	
Control Test		0.078	23:59	
Air Blank		0.000	23:59	
Subject Sample	# 1		23:59	Subject Test Refused
Breath Volume		0.000	23:59	
Air Blank		0.000	00:00	
Control Test		0.078	00:00	
Air Blank		0.000	00:01	
Diagnostics Check		OK	00:01	

Data Download Date/Time: 04/23/2019 16:18

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: MUNOZ

DATE/TIME OF TEST: 04/03/2019 04:13:51

Diagnostics Check		OK	04:10	
Air Blank		0.000	04:11	
Control Test		0.078	04:11	
Air Blank		0.000	04:12	
Subject Sample	# 1		04:12	Subject Test Refused
Breath Volume		0.000	04:12	
Air Blank		0.000	04:12	
Control Test		0.078	04:13	
Air Blank		0.000	04:13	
Diagnostics Check		OK	04:13	

Data Download Date/Time: 04/23/2019 16:18

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPURLOCK,ROBERT,V
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SPURLOCK,ROBERT
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: ADMIN
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 04/03/2019 19:34:47

Diagnostics Check		OK	19:29
Air Blank		0.000	19:29
Control Test		0.076	19:29
Air Blank		0.000	19:30
Subject Sample	# 1	0.000	19:30
Breath Volume		2.230	19:30
Air Blank		0.000	19:31
Air Blank		0.000	19:33
Subject Sample	# 2	0.000	19:33
Breath Volume		2.191	19:33
Air Blank		0.000	19:33
Control Test		0.077	19:34
Air Blank		0.000	19:34
Diagnostics Check		OK	19:34

Data Download Date/Time: 04/23/2019 16:18

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 19:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:01

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 04/05/2019 20:32:13

Diagnostics Check		OK	20:25
Air Blank		0.000	20:25
Control Test		0.077	20:25
Air Blank		0.000	20:26
Subject Sample	# 1	0.155	20:27
Breath Volume		1.675	20:27
Air Blank		0.000	20:28
Air Blank		0.000	20:29
Subject Sample	# 2	0.151	20:30
Breath Volume		2.304	20:30
Air Blank		0.000	20:31
Control Test		0.075	20:31
Air Blank		0.000	20:32
Diagnostics Check		OK	20:32

Data Download Date/Time: 04/23/2019 16:18

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: BOUREMANA

DATE/TIME OF TEST: 04/05/2019 23:38:36

Diagnostics Check		OK	23:32
Air Blank		0.000	23:32
Control Test		0.077	23:33
Air Blank		0.000	23:33
Subject Sample	# 1	0.000	23:34
Breath Volume		2.292	23:34
Air Blank		0.000	23:34
Air Blank		0.000	23:36
Subject Sample	# 2	0.000	23:37
Breath Volume		2.183	23:37
Air Blank		0.000	23:37
Control Test		0.075	23:37
Air Blank		0.000	23:38
Diagnostics Check		OK	23:38

Data Download Date/Time: 04/23/2019 16:18

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 04/06/2019 02:37:20

Diagnostics Check		OK	02:28
Air Blank		0.000	02:29
Control Test		0.077	02:29
Air Blank		0.000	02:30
Subject Sample	# 1	0.193	02:31
Breath Volume		1.527	02:31
Air Blank		0.000	02:32
Air Blank		0.000	02:33
Subject Sample	# 2	0.184	02:35
Breath Volume		1.886	02:35
Air Blank		0.000	02:36
Control Test		0.076	02:36
Air Blank		0.000	02:37
Diagnostics Check		OK	02:37

Data Download Date/Time: 04/23/2019 16:19

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 19:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: WORKMAN

DATE/TIME OF TEST: 04/06/2019 20:45:04

Diagnostics Check		OK	20:38
Air Blank		0.000	20:38
Control Test		0.077	20:38
Air Blank		0.000	20:39
Subject Sample	# 1	0.142	20:40
Breath Volume		2.535	20:40
Air Blank		0.000	20:41
Air Blank		0.000	20:42
Subject Sample	# 2	0.136	20:43
Breath Volume		2.406	20:43
Air Blank		0.000	20:44
Control Test		0.076	20:44
Air Blank		0.000	20:44
Diagnostics Check		OK	20:45

Data Download Date/Time: 04/23/2019 16:19

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:11

SUBJECT'S LAST NAME: AGUIRRE LOPEZ]

DATE/TIME OF TEST: 04/06/2019 22:38:10

Diagnostics Check		OK	22:35	
Air Blank		0.000	22:35	
Control Test		0.078	22:36	
Air Blank		0.000	22:36	
Subject Sample	# 1		22:36	Subject Test Refused
Breath Volume		0.000	22:36	
Air Blank		0.000	22:37	
Control Test		0.078	22:37	
Air Blank		0.000	22:38	
Diagnostics Check		OK	22:38	

Data Download Date/Time: 04/23/2019 16:19

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: HERNANDEZ VALLE

DATE/TIME OF TEST: 04/07/2019 03:51:16

Diagnostics Check		OK	03:44
Air Blank		0.000	03:44
Control Test		0.077	03:45
Air Blank		0.000	03:45
Subject Sample	# 1	0.149	03:46
Breath Volume		3.761	03:46
Air Blank		0.000	03:47
Air Blank		0.000	03:49
Subject Sample	# 2	0.148	03:49
Breath Volume		3.265	03:49
Air Blank		0.000	03:50
Control Test		0.077	03:50
Air Blank		0.000	03:51
Diagnostics Check		OK	03:51

Data Download Date/Time: 04/23/2019 16:19

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMBRIATI,NICHOLAS,J
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: AMBRIATI,NICHOLAS
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 11:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:01

SUBJECT'S LAST NAME: ZIGLAR

DATE/TIME OF TEST: 04/08/2019 12:37:49

Diagnostics Check		OK	12:31
Air Blank		0.000	12:31
Control Test		0.077	12:32
Air Blank		0.000	12:32
Subject Sample	# 1	0.182	12:33
Breath Volume		3.003	12:33
Air Blank		0.000	12:33
Air Blank		0.000	12:35
Subject Sample	# 2	0.187	12:36
Breath Volume		3.226	12:36
Air Blank		0.000	12:36
Control Test		0.076	12:37
Air Blank		0.000	12:37
Diagnostics Check		OK	12:37

Data Download Date/Time: 04/23/2019 16:19

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,KATRYNA,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TAYLOR,KATRYNA
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: ESPINAL-ROSALES

DATE/TIME OF TEST: 04/10/2019 03:23:18

Diagnostics Check		OK	03:16
Air Blank		0.000	03:17
Control Test		0.077	03:17
Air Blank		0.000	03:18
Subject Sample	# 1	0.204	03:18
Breath Volume		1.957	03:18
Air Blank		0.000	03:19
Air Blank		0.000	03:21
Subject Sample	# 2	0.197	03:21
Breath Volume		1.921	03:21
Air Blank		0.000	03:22
Control Test		0.076	03:22
Air Blank		0.000	03:23
Diagnostics Check		OK	03:23

Data Download Date/Time: 04/23/2019 16:19

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: LEAL LEAL

DATE/TIME OF TEST: 04/11/2019 04:32:22

Diagnostics Check		OK	04:24
Air Blank		0.000	04:25
Control Test		0.077	04:25
Air Blank		0.000	04:25
Subject Sample	# 1	0.185	04:27
Breath Volume		1.207	04:27
Air Blank		0.000	04:28
Air Blank		0.000	04:30
Subject Sample	# 2	0.193	04:30
Breath Volume		1.667	04:30
Air Blank		0.000	04:31
Control Test		0.075	04:31
Air Blank		0.000	04:32
Diagnostics Check		OK	04:32

Data Download Date/Time: 04/23/2019 16:19

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: ANDERSEN

DATE/TIME OF TEST: 04/15/2019 23:56:22

Diagnostics Check		OK	23:46
Air Blank		0.000	23:46
Control Test		0.078	23:46
Air Blank		0.000	23:47
Subject Sample	# 1	0.172	23:48
Breath Volume		2.976	23:48
Air Blank		0.000	23:49
Air Blank		0.000	23:50
Subject Sample	# 2	0.144	23:51
Breath Volume		1.839	23:51
Air Blank		0.000	23:52
Air Blank		0.000	23:54
Subject Sample	# 3	0.155	23:54
Breath Volume		1.894	23:54
Air Blank		0.000	23:55
Control Test		0.078	23:55
Air Blank		0.000	23:56
Diagnostics Check		OK	23:56

Data Download Date/Time: 04/23/2019 16:19

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: LOVE

DATE/TIME OF TEST: 04/16/2019 00:28:03

Diagnostics Check		OK	00:21
Air Blank		0.000	00:21
Control Test		0.078	00:22
Air Blank		0.000	00:22
Subject Sample	# 1	0.164	00:23
Breath Volume		3.988	00:23
Air Blank		0.000	00:23
Air Blank		0.000	00:25
Subject Sample	# 2	0.162	00:26
Breath Volume		2.082	00:26
Air Blank		0.000	00:27
Control Test		0.077	00:27
Air Blank		0.000	00:27
Diagnostics Check		OK	00:28

Data Download Date/Time: 04/23/2019 16:19

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: GRATZER

DATE/TIME OF TEST: 04/16/2019 04:04:09

Diagnostics Check		OK	04:01	
Air Blank		0.000	04:01	
Control Test		0.078	04:02	
Air Blank		0.000	04:02	
Subject Sample	# 1		04:02	Subject Test Refused
Breath Volume		0.000	04:02	
Air Blank		0.000	04:03	
Control Test		0.078	04:03	
Air Blank		0.000	04:04	
Diagnostics Check		OK	04:04	

Data Download Date/Time: 04/23/2019 16:19

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: PINKNEY

DATE/TIME OF TEST: 04/17/2019 02:01:07

Diagnostics Check		OK	01:54
Air Blank		0.000	01:54
Control Test		0.078	01:55
Air Blank		0.000	01:55
Subject Sample	# 1	0.126	01:56
Breath Volume		4.070	01:56
Air Blank		0.000	01:56
Air Blank		0.000	01:58
Subject Sample	# 2	0.131	01:59
Breath Volume		4.597	01:59
Air Blank		0.000	02:00
Control Test		0.077	02:00
Air Blank		0.000	02:01
Diagnostics Check		OK	02:01

Data Download Date/Time: 04/23/2019 16:19

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: AGUILAR PAREDES

DATE/TIME OF TEST: 04/20/2019 01:36:45

Diagnostics Check		OK	01:33	
Air Blank		0.000	01:33	
Control Test		0.077	01:33	
Air Blank		0.000	01:34	
Subject Sample	# 1		01:35	Subject Test Refused
Breath Volume		0.000	01:35	
Air Blank		0.000	01:35	
Control Test		0.078	01:36	
Air Blank		0.000	01:36	
Diagnostics Check		OK	01:36	

Data Download Date/Time: 04/23/2019 16:19

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:47

SUBJECT'S LAST NAME: GARCIA GOMEZ

DATE/TIME OF TEST: 04/20/2019 22:14:35

Diagnostics Check		OK	22:07
Air Blank		0.000	22:08
Control Test		0.077	22:08
Air Blank		0.000	22:09
Subject Sample	# 1	0.170	22:09
Breath Volume		1.804	22:09
Air Blank		0.000	22:10
Air Blank		0.000	22:12
Subject Sample	# 2	0.170	22:12
Breath Volume		1.546	22:12
Air Blank		0.000	22:13
Control Test		0.076	22:13
Air Blank		0.000	22:14
Diagnostics Check		OK	22:14

Data Download Date/Time: 04/23/2019 16:19

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: MANNING

DATE/TIME OF TEST: 04/20/2019 23:17:05

Diagnostics Check		OK	23:09
Air Blank		0.000	23:09
Control Test		0.077	23:09
Air Blank		0.000	23:10
Subject Sample	# 1	0.110	23:11
Breath Volume		1.851	23:11
Air Blank		0.000	23:12
Air Blank		0.000	23:14
Subject Sample	# 2	0.110	23:15
Breath Volume		1.652	23:15
Air Blank		0.000	23:16
Control Test		0.077	23:16
Air Blank		0.000	23:17
Diagnostics Check		OK	23:17

Data Download Date/Time: 04/23/2019 16:19

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANO,ADRIEL,
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: SALAZAR,L
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 04/16/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 04/18/2019 00:32:17

Diagnostics Check		OK	00:23
Air Blank		0.000	00:24
Control Test		0.077	00:24
Air Blank		0.000	00:25
Subject Sample	# 1	0.195	00:27
Breath Volume		2.742	00:27
Air Blank		0.000	00:28
Air Blank		0.000	00:30
Subject Sample	# 2	0.190	00:30
Breath Volume		2.820	00:30
Air Blank		0.000	00:31
Control Test		0.075	00:31
Air Blank		0.000	00:32
Diagnostics Check		OK	00:32

Data Download Date/Time: 05/23/2019 11:04

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANO,ADRIEL,
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: GONZALEZ,EMANUEL
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 04/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: HERCULES LANDAVERDE

DATE/TIME OF TEST: 04/19/2019 22:45:10

Diagnostics Check		OK	22:42	
Air Blank		0.000	22:42	
Control Test		0.076	22:43	
Air Blank		0.000	22:43	
Subject Sample	# 1		22:43	Subject Test Refused
Breath Volume		0.000	22:43	
Air Blank		0.000	22:44	
Control Test		0.074	22:44	Control Outside Tolerance
Air Blank		0.000	22:45	

Data Download Date/Time: 05/23/2019 11:04

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANO,ADRIEL,
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: VILORIO,TAYRON
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 04/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 04/19/2019 23:50:36

Diagnostics Check		OK	23:44
Air Blank		0.000	23:44
Control Test		0.077	23:45
Air Blank		0.000	23:45
Subject Sample	# 1	0.181	23:46
Breath Volume		1.187	23:46
Air Blank		0.000	23:46
Air Blank		0.000	23:48
Subject Sample	# 2	0.178	23:49
Breath Volume		1.824	23:49
Air Blank		0.000	23:49
Control Test		0.075	23:49
Air Blank		0.000	23:50
Diagnostics Check		OK	23:50

Data Download Date/Time: 05/23/2019 11:04

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUNOZ,A,
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: QUINTEROS,M
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 04/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 13:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:05

SUBJECT'S LAST NAME: CABRERA

DATE/TIME OF TEST: 04/21/2019 14:34:25

Diagnostics Check		OK	14:31	
Air Blank		0.000	14:32	
Control Test		0.077	14:32	
Air Blank		0.000	14:32	
Subject Sample	# 1		14:33	Subject Test Refused
Breath Volume		0.000	14:33	
Air Blank		0.000	14:33	
Control Test		0.077	14:33	
Air Blank		0.000	14:34	
Diagnostics Check		OK	14:34	

Data Download Date/Time: 05/23/2019 11:04

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANCHEZ,RAUL,
BREATH TEST OPERATOR AGENCY: Bal Harbour Vil PD
ARRESTING OFFICER: LAHARA,L
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 04/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 14:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:53

SUBJECT'S LAST NAME: MARCHAN

DATE/TIME OF TEST: 04/28/2019 15:35:32

Diagnostics Check		OK	15:28
Air Blank		0.000	15:29
Control Test		0.077	15:29
Air Blank		0.000	15:29
Subject Sample	# 1	0.237	15:30
Breath Volume		1.156	15:30
Air Blank		0.000	15:31
Air Blank		0.000	15:33
Subject Sample	# 2	0.235	15:33
Breath Volume		1.710	15:33
Air Blank		0.000	15:34
Control Test		0.075	15:34
Air Blank		0.000	15:35
Diagnostics Check		OK	15:35

Data Download Date/Time: 05/23/2019 11:04

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,ALEXIS,
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: ELIAS,FELIX
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 04/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: ARIAS

DATE/TIME OF TEST: 04/30/2019 23:39:34

Diagnostics Check		OK	23:36	
Air Blank		0.000	23:37	
Control Test		0.077	23:37	
Air Blank		0.000	23:38	
Subject Sample	# 1		23:38	Subject Test Refused
Breath Volume		0.000	23:38	
Air Blank		0.000	23:38	
Control Test		0.077	23:38	
Air Blank		0.000	23:39	
Diagnostics Check		OK	23:39	

Data Download Date/Time: 05/23/2019 11:04

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002464

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARROJAS,GABRIEL,L
BREATH TEST OPERATOR AGENCY: HIALEAH PD
ARRESTING OFFICER: MEJIA,BOBBY
ARRESTING OFFICER AGENCY: HIALEAH PD
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: MENDOZA

DATE/TIME OF TEST: 04/06/2019 03:07:30

Diagnostics Check		OK	03:04	
Air Blank		0.000	03:05	
Control Test		0.077	03:05	
Air Blank		0.000	03:05	
Subject Sample	# 1		03:06	Subject Test Refused
Breath Volume		0.000	03:06	
Air Blank		0.000	03:06	
Control Test		0.078	03:06	
Air Blank		0.000	03:07	
Diagnostics Check		OK	03:07	

Data Download Date/Time: 04/16/2019 11:13

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002464

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAGOS,GUILLERMO,E
BREATH TEST OPERATOR AGENCY: MEDLEY PD
ARRESTING OFFICER: NICKERSON,BRIAN
ARRESTING OFFICER AGENCY: MIAMI SPRINGS PD
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 12:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:20

SUBJECT'S LAST NAME: SHELOW

DATE/TIME OF TEST: 04/14/2019 14:08:35

Diagnostics Check		OK	14:05	
Air Blank		0.000	14:06	
Control Test		0.078	14:06	
Air Blank		0.000	14:06	
Subject Sample	# 1		14:07	Subject Test Refused
Breath Volume		0.000	14:07	
Air Blank		0.000	14:07	
Control Test		0.077	14:07	
Air Blank		0.000	14:08	
Diagnostics Check		OK	14:08	

Data Download Date/Time: 04/16/2019 11:13

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002566

INSTRUMENT REGISTERED TO: SURFSIDE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAVAS,DAVID,E
BREATH TEST OPERATOR AGENCY: BAL HARBOUR VIL PD
ARRESTING OFFICER: BLANCO,ERICA
ARRESTING OFFICER AGENCY: BAL HARBOUR VIL PD
LAST AGENCY INSPECTION: 03/27/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: MAXIMOVA

DATE/TIME OF TEST: 04/22/2019 02:50:48

Diagnostics Check	Fail	02:50	DSP Test
Air Blank		02:50	Ambient Fail

Data Download Date/Time: 06/27/2019 10:21

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002566

INSTRUMENT REGISTERED TO: SURFSIDE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAVAS,DAVID,E
BREATH TEST OPERATOR AGENCY: BAL HARBOUR VIL PD
ARRESTING OFFICER: BLANCO,ERICA
ARRESTING OFFICER AGENCY: BAL HARBOUR VIL PD
LAST AGENCY INSPECTION: 03/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: MAXIMOVA

DATE/TIME OF TEST: 04/22/2019 02:53:34

Diagnostics Check	Fail	02:53	DSP Test
Air Blank		02:53	Ambient Fail

Data Download Date/Time: 06/27/2019 10:21

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002987

INSTRUMENT REGISTERED TO: MIAMI GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OMALLEY,BRIAN,P
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: GEBBIA,SALVATORE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A2
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 04/13/2019 22:06:14

Diagnostics Check		OK	22:03	
Air Blank		0.000	22:03	
Control Test		0.077	22:04	
Air Blank		0.000	22:04	
Subject Sample	# 1		22:04	Subject Test Refused
Breath Volume		0.000	22:04	
Air Blank		0.000	22:05	
Control Test		0.078	22:05	
Air Blank		0.000	22:06	
Diagnostics Check		OK	22:06	

Data Download Date/Time: 04/30/2019 15:29

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002988

INSTRUMENT REGISTERED TO: MIAMI GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HODGES,TITUS,R
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: HODGES,TITUS
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: PRENDERGAST

DATE/TIME OF TEST: 04/06/2019 04:07:13

Diagnostics Check		OK	03:59
Air Blank		0.000	04:00
Control Test		0.080	04:00
Air Blank		0.000	04:00
Subject Sample	# 1	0.146	04:02
Breath Volume		1.171	04:02
Air Blank		0.000	04:02
Air Blank		0.000	04:04
Subject Sample	# 2	0.153	04:05
Breath Volume		1.359	04:05
Air Blank		0.000	04:06
Control Test		0.079	04:06
Air Blank		0.000	04:07
Diagnostics Check		OK	04:07

Data Download Date/Time: 04/30/2019 15:36

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002988

INSTRUMENT REGISTERED TO: MIAMI GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HODGES,TITUS,R
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: BLANCO,BRAIN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: ARES

DATE/TIME OF TEST: 04/07/2019 05:13:47

Diagnostics Check		OK	05:04	
Air Blank		0.000	05:05	
Control Test		0.080	05:05	
Air Blank		0.000	05:06	
Subject Sample	# 1		05:09	No Sample Provided
Breath Volume		0.000	05:09	
Air Blank		0.000	05:09	
Air Blank		0.000	05:11	
Subject Sample	# 2		05:12	Subject Test Refused
Breath Volume		0.000	05:12	
Air Blank		0.000	05:12	
Control Test		0.080	05:13	
Air Blank		0.000	05:13	
Diagnostics Check		OK	05:13	

Data Download Date/Time: 04/30/2019 15:36

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002988

INSTRUMENT REGISTERED TO: MIAMI GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OMALLEY,BRIAN,P
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: OMALLEY,BRIAN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A2
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 13:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:55

SUBJECT'S LAST NAME: MCRAE

DATE/TIME OF TEST: 04/09/2019 14:50:57

Diagnostics Check		OK	14:44
Air Blank		0.000	14:45
Control Test		0.080	14:45
Air Blank		0.000	14:46
Subject Sample	# 1	0.236	14:46
Breath Volume		1.882	14:46
Air Blank		0.000	14:47
Air Blank		0.000	14:49
Subject Sample	# 2	0.234	14:49
Breath Volume		1.214	14:49
Air Blank		0.000	14:50
Control Test		0.080	14:50
Air Blank		0.000	14:50
Diagnostics Check		OK	14:50

Data Download Date/Time: 04/30/2019 15:36

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002988

INSTRUMENT REGISTERED TO: MIAMI GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAMAYO,KEVIN,X
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: TAMAYO,KEVIN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A2
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 06:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:50

SUBJECT'S LAST NAME: HUDSON

DATE/TIME OF TEST: 04/10/2019 07:18:40

Diagnostics Check		OK	07:14	
Air Blank		0.000	07:14	
Control Test		0.079	07:14	
Air Blank		0.000	07:15	
Subject Sample	# 1		07:18	Subject Test Refused
Breath Volume		0.066	07:18	
Air Blank			07:18	RFI Detect
Air Blank			07:18	RFI Detect

Data Download Date/Time: 04/30/2019 15:36

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002988

INSTRUMENT REGISTERED TO: MIAMI GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HODGES,TITUS,R
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: BLANCO,BRIAN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:19

SUBJECT'S LAST NAME: HYPPOLITE

DATE/TIME OF TEST: 04/20/2019 05:48:51

Diagnostics Check		OK	05:40	
Air Blank		0.000	05:40	
Control Test		0.080	05:40	
Air Blank		0.000	05:41	
Subject Sample	# 1		05:44	No Sample Provided
Breath Volume		0.000	05:44	
Air Blank		0.000	05:44	
Air Blank		0.000	05:46	
Subject Sample	# 2		05:47	Subject Test Refused
Breath Volume		0.000	05:47	
Air Blank		0.000	05:47	
Control Test		0.080	05:48	
Air Blank		0.000	05:48	
Diagnostics Check		OK	05:48	

Data Download Date/Time: 04/30/2019 15:36

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002988

INSTRUMENT REGISTERED TO: MIAMI GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HODGES,TITUS,R
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: HODGES,TITUS
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: CHRISTIAN

DATE/TIME OF TEST: 04/26/2019 02:01:24

Diagnostics Check		OK	01:50
Air Blank		0.000	01:51
Control Test		0.080	01:51
Air Blank		0.000	01:52
Subject Sample	# 1	0.166	01:54
Breath Volume		1.488	01:54
Air Blank		0.000	01:54
Air Blank		0.000	01:56
Subject Sample	# 2	0.181	01:59
Breath Volume		1.167	01:59
Air Blank		0.000	02:00
Control Test		0.080	02:00
Air Blank		0.000	02:01
Diagnostics Check		OK	02:01

Data Download Date/Time: 04/30/2019 15:36

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROUCH,CLIFFORD,D
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: CROUCH,CLIFFORD
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 03/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: WIGGINS

DATE/TIME OF TEST: 04/04/2019 01:51:48

Diagnostics Check		OK	01:45
Air Blank		0.000	01:45
Control Test		0.081	01:45
Air Blank		0.000	01:46
Subject Sample	# 1	0.136	01:47
Breath Volume		4.046	01:47
Air Blank		0.000	01:48
Air Blank		0.000	01:49
Subject Sample	# 2	0.135	01:50
Breath Volume		3.867	01:50
Air Blank		0.000	01:50
Control Test		0.080	01:51
Air Blank		0.000	01:51
Diagnostics Check		OK	01:51

Data Download Date/Time: 05/02/2019 10:59

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,A
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,GARY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 03/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: FELIX

DATE/TIME OF TEST: 04/24/2019 01:41:14

Diagnostics Check		OK	01:30
Air Blank		0.000	01:31
Control Test		0.082	01:31
Air Blank		0.000	01:32
Subject Sample	# 1	0.275	01:34
Breath Volume		1.453	01:34
Air Blank		0.000	01:35
Air Blank		0.000	01:37
Subject Sample	# 2	0.263	01:39
Breath Volume		1.746	01:39
Air Blank		0.000	01:40
Control Test		0.077	01:40
Air Blank		0.000	01:41
Diagnostics Check		OK	01:41

Data Download Date/Time: 05/02/2019 10:59

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,GARY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 03/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: INGRAM

DATE/TIME OF TEST: 04/24/2019 21:53:39

Diagnostics Check		OK	21:47
Air Blank		0.000	21:47
Control Test		0.079	21:47
Air Blank		0.000	21:48
Subject Sample	# 1	0.163	21:48
Breath Volume		3.339	21:48
Air Blank		0.000	21:49
Air Blank		0.000	21:51
Subject Sample	# 2	0.166	21:51
Breath Volume		3.187	21:51
Air Blank		0.000	21:52
Control Test		0.079	21:53
Air Blank		0.000	21:53
Diagnostics Check		OK	21:53

Data Download Date/Time: 05/02/2019 10:59

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003196

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNTON,KENNETH,W
BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)
ARRESTING OFFICER: THORNTON,KENNETH
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 04/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: BUI
ARREST TIME: 20:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: TORRY

DATE/TIME OF TEST: 04/28/2019 21:41:49

Diagnostics Check		OK	21:35
Air Blank		0.000	21:36
Control Test		0.078	21:36
Air Blank		0.000	21:36
Subject Sample	# 1	0.128	21:37
Breath Volume		1.621	21:37
Air Blank		0.000	21:38
Air Blank		0.000	21:39
Subject Sample	# 2	0.126	21:40
Breath Volume		1.953	21:40
Air Blank		0.000	21:40
Control Test		0.077	21:41
Air Blank		0.000	21:41
Diagnostics Check		OK	21:41

Data Download Date/Time: 05/17/2019 15:01

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOSA,JONATHAN,
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: SOSA,JONATHAN
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 03/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 04/06/2019 23:12:42

Diagnostics Check		OK	23:04
Air Blank		0.000	23:05
Control Test		0.080	23:05
Air Blank		0.000	23:06
Subject Sample	# 1	0.218	23:08
Breath Volume		2.359	23:08
Air Blank		0.000	23:09
Air Blank		0.000	23:10
Subject Sample	# 2	0.218	23:10
Breath Volume		1.679	23:10
Air Blank		0.000	23:11
Control Test		0.077	23:12
Air Blank		0.000	23:12
Diagnostics Check		OK	23:12

Data Download Date/Time: 04/26/2019 11:54

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,BARRY,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: NIXON,BARRY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 04/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 15:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:55

SUBJECT'S LAST NAME: VANSICKLE

DATE/TIME OF TEST: 04/24/2019 16:22:17

Diagnostics Check		OK	16:16
Air Blank		0.000	16:16
Control Test		0.080	16:17
Air Blank		0.000	16:17
Subject Sample	# 1	0.219	16:17
Breath Volume		1.898	16:17
Air Blank		0.000	16:18
Air Blank		0.000	16:20
Subject Sample	# 2	0.227	16:20
Breath Volume		1.105	16:20
Air Blank		0.000	16:21
Control Test		0.080	16:21
Air Blank		0.000	16:22
Diagnostics Check		OK	16:22

Data Download Date/Time: 05/17/2019 11:37

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSENBLOOM,MATTHEW,S
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: HAMMELMAN,TYLER
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 04/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: COLEY

DATE/TIME OF TEST: 04/25/2019 05:03:40

Diagnostics Check		OK	04:55	
Air Blank		0.000	04:55	
Control Test		0.080	04:55	
Air Blank		0.000	04:56	
Subject Sample	# 1	0.054	04:58	
Breath Volume		1.406	04:58	
Air Blank		0.000	04:58	
Air Blank		0.000	05:00	
Subject Sample	# 2	0.064	05:02	
Breath Volume		2.375	05:02	
Air Blank		0.000	05:02	
Control Test		0.082	05:02	
Air Blank			05:03	RFI Detect
Air Blank		0.000	05:03	

Data Download Date/Time: 05/17/2019 11:37

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSENBLOOM,MATTHEW,S
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: HAMMELMAN,TYLER
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 04/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: COLEY

DATE/TIME OF TEST: 04/25/2019 05:06:30

Diagnostics Check	OK	05:05	
Air Blank		05:06	RFI Detect
Air Blank	0.000	05:06	

Data Download Date/Time: 05/17/2019 11:37

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSENBLOOM,MATTHEW,S
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: HAMMELMAN,TYLER
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 04/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: COLEY

DATE/TIME OF TEST: 04/25/2019 05:15:19

Diagnostics Check		OK	05:08
Air Blank		0.000	05:08
Control Test		0.082	05:08
Air Blank		0.000	05:09
Subject Sample	# 1	0.063	05:10
Breath Volume		3.039	05:10
Air Blank		0.000	05:11
Air Blank		0.000	05:13
Subject Sample	# 2	0.061	05:13
Breath Volume		2.812	05:13
Air Blank		0.000	05:14
Control Test		0.080	05:14
Air Blank		0.000	05:15
Diagnostics Check		OK	05:15

Data Download Date/Time: 05/17/2019 11:37

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: MUNGUIA-LOPEZ

DATE/TIME OF TEST: 04/05/2019 21:31:49

Diagnostics Check		OK	21:25
Air Blank		0.000	21:25
Control Test		0.079	21:25
Air Blank		0.000	21:26
Subject Sample	# 1	0.277	21:27
Breath Volume		2.929	21:27
Air Blank		0.000	21:27
Air Blank		0.000	21:29
Subject Sample	# 2	0.271	21:30
Breath Volume		3.125	21:30
Air Blank		0.000	21:30
Control Test		0.079	21:31
Air Blank		0.000	21:31
Diagnostics Check		OK	21:31

Data Download Date/Time: 05/06/2019 14:34

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DRISCOLL,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: RIOS

DATE/TIME OF TEST: 04/06/2019 04:24:21

Diagnostics Check		OK	04:18
Air Blank		0.000	04:18
Control Test		0.079	04:19
Air Blank		0.000	04:19
Subject Sample	# 1	0.068	04:20
Breath Volume		2.167	04:20
Air Blank		0.000	04:20
Air Blank		0.000	04:22
Subject Sample	# 2	0.069	04:22
Breath Volume		2.890	04:22
Air Blank		0.000	04:23
Control Test		0.079	04:23
Air Blank		0.000	04:24
Diagnostics Check		OK	04:24

Data Download Date/Time: 05/06/2019 14:34

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 04/06/2019 05:02:14

Diagnostics Check		OK	04:49	
Air Blank		0.000	04:49	
Control Test		0.079	04:50	
Air Blank		0.000	04:50	
Subject Sample	# 1	0.177	04:53	Volume Not Met
Breath Volume		1.097	04:53	
Air Blank		0.000	04:54	
Air Blank		0.000	04:56	
Subject Sample	# 2	0.170	04:57	
Breath Volume		2.003	04:57	
Air Blank		0.000	04:57	
Air Blank		0.000	04:59	
Subject Sample	# 3	0.172	05:00	
Breath Volume		1.425	05:00	
Air Blank		0.000	05:01	
Control Test		0.078	05:01	
Air Blank		0.000	05:02	
Diagnostics Check		OK	05:02	

Data Download Date/Time: 05/06/2019 14:34

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES AVERY,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:46

SUBJECT'S LAST NAME: RICHARDS

DATE/TIME OF TEST: 04/06/2019 23:12:45

Diagnostics Check		OK	23:06
Air Blank		0.000	23:07
Control Test		0.079	23:07
Air Blank		0.000	23:08
Subject Sample	# 1	0.157	23:08
Breath Volume		3.703	23:08
Air Blank		0.000	23:09
Air Blank		0.000	23:10
Subject Sample	# 2	0.156	23:11
Breath Volume		2.628	23:11
Air Blank		0.000	23:11
Control Test		0.078	23:12
Air Blank		0.000	23:12
Diagnostics Check		OK	23:12

Data Download Date/Time: 05/06/2019 14:34

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES AVERY,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SCHAEFER,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: BARRERA SANTANA

DATE/TIME OF TEST: 04/07/2019 01:36:24

Diagnostics Check		OK	01:30
Air Blank		0.000	01:30
Control Test		0.079	01:31
Air Blank		0.000	01:31
Subject Sample	# 1	0.159	01:31
Breath Volume		1.289	01:31
Air Blank		0.000	01:32
Air Blank		0.000	01:34
Subject Sample	# 2	0.162	01:34
Breath Volume		1.855	01:34
Air Blank		0.000	01:35
Control Test		0.078	01:35
Air Blank		0.000	01:36
Diagnostics Check		OK	01:36

Data Download Date/Time: 05/06/2019 14:35

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: MURILLO CONTRERAS

DATE/TIME OF TEST: 04/07/2019 03:57:31

Diagnostics Check		OK	03:50
Air Blank		0.000	03:50
Control Test		0.080	03:51
Air Blank		0.000	03:51
Subject Sample	# 1	0.226	03:52
Breath Volume		1.707	03:52
Air Blank		0.000	03:53
Air Blank		0.000	03:55
Subject Sample	# 2	0.223	03:55
Breath Volume		1.730	03:55
Air Blank		0.000	03:56
Control Test		0.078	03:56
Air Blank		0.000	03:57
Diagnostics Check		OK	03:57

Data Download Date/Time: 05/06/2019 14:35

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES AVERY,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VASALLO,Y
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 04/07/2019 22:14:43

Diagnostics Check		OK	22:07
Air Blank		0.000	22:07
Control Test		0.079	22:07
Air Blank		0.000	22:08
Subject Sample	# 1	0.235	22:09
Breath Volume		1.843	22:09
Air Blank		0.000	22:10
Air Blank		0.000	22:11
Subject Sample	# 2	0.216	22:13
Breath Volume		1.226	22:13
Air Blank		0.000	22:13
Control Test		0.078	22:14
Air Blank		0.000	22:14
Diagnostics Check		OK	22:14

Data Download Date/Time: 05/06/2019 14:35

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: COSTA,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:11

SUBJECT'S LAST NAME: BECKTON

DATE/TIME OF TEST: 04/07/2019 22:38:10

Diagnostics Check		OK	22:31
Air Blank		0.000	22:32
Control Test		0.078	22:32
Air Blank		0.000	22:33
Subject Sample	# 1	0.226	22:33
Breath Volume		2.523	22:33
Air Blank		0.000	22:34
Air Blank		0.000	22:35
Subject Sample	# 2	0.226	22:36
Breath Volume		2.921	22:36
Air Blank		0.000	22:37
Control Test		0.079	22:37
Air Blank		0.000	22:38
Diagnostics Check		OK	22:38

Data Download Date/Time: 05/06/2019 14:35

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: FISH

DATE/TIME OF TEST: 04/10/2019 00:02:48

Diagnostics Check		OK	23:56
Air Blank		0.000	23:56
Control Test		0.079	23:57
Air Blank		0.000	23:57
Subject Sample	# 2	0.000	23:58
Breath Volume		4.339	23:58
Air Blank		0.000	23:58
Air Blank		0.000	00:00
Subject Sample	# 1	0.000	00:01
Breath Volume		4.433	00:01
Air Blank		0.000	00:01
Control Test		0.079	00:02
Air Blank		0.000	00:02
Diagnostics Check		OK	00:02

Data Download Date/Time: 05/06/2019 14:35

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: LEITER

DATE/TIME OF TEST: 04/11/2019 01:16:47

Diagnostics Check		OK	01:10
Air Blank		0.000	01:10
Control Test		0.080	01:11
Air Blank		0.000	01:11
Subject Sample	# 1	0.179	01:12
Breath Volume		2.234	01:12
Air Blank		0.000	01:12
Air Blank		0.000	01:14
Subject Sample	# 2	0.180	01:15
Breath Volume		1.828	01:15
Air Blank		0.000	01:15
Control Test		0.078	01:16
Air Blank		0.000	01:16
Diagnostics Check		OK	01:16

Data Download Date/Time: 05/06/2019 14:35

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ARCHAMBEAU,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: JOY

DATE/TIME OF TEST: 04/11/2019 19:41:45

Diagnostics Check		OK	19:35
Air Blank		0.000	19:35
Control Test		0.079	19:36
Air Blank		0.000	19:36
Subject Sample	# 1	0.116	19:37
Breath Volume		2.199	19:37
Air Blank		0.000	19:37
Air Blank		0.000	19:39
Subject Sample	# 2	0.115	19:40
Breath Volume		2.410	19:40
Air Blank		0.000	19:40
Control Test		0.079	19:41
Air Blank		0.000	19:41
Diagnostics Check		OK	19:41

Data Download Date/Time: 05/06/2019 14:35

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIMPKINS,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 05:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:46

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 04/12/2019 06:13:03

Diagnostics Check		OK	06:06
Air Blank		0.000	06:06
Control Test		0.079	06:07
Air Blank		0.000	06:07
Subject Sample	# 1	0.230	06:08
Breath Volume		2.617	06:08
Air Blank		0.000	06:09
Air Blank		0.000	06:10
Subject Sample	# 2	0.219	06:11
Breath Volume		2.285	06:11
Air Blank		0.000	06:12
Control Test		0.079	06:12
Air Blank		0.000	06:12
Diagnostics Check		OK	06:13

Data Download Date/Time: 05/06/2019 14:35

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,TIMOTHY,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEE,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:10

SUBJECT'S LAST NAME: BARNHART

DATE/TIME OF TEST: 04/12/2019 17:01:36

Diagnostics Check		OK	16:55
Air Blank		0.000	16:55
Control Test		0.079	16:55
Air Blank		0.000	16:56
Subject Sample	# 1	0.205	16:57
Breath Volume		1.824	16:57
Air Blank		0.000	16:57
Air Blank		0.000	16:59
Subject Sample	# 2	0.201	17:00
Breath Volume		1.816	17:00
Air Blank		0.000	17:00
Control Test		0.078	17:00
Air Blank		0.000	17:01
Diagnostics Check		OK	17:01

Data Download Date/Time: 05/06/2019 14:35

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: MEDINA

DATE/TIME OF TEST: 04/12/2019 22:24:00

Diagnostics Check		OK	22:17
Air Blank		0.000	22:18
Control Test		0.079	22:18
Air Blank		0.000	22:18
Subject Sample	# 1	0.191	22:19
Breath Volume		2.097	22:19
Air Blank		0.000	22:19
Air Blank		0.000	22:21
Subject Sample	# 2	0.188	22:22
Breath Volume		2.570	22:22
Air Blank		0.000	22:23
Control Test		0.078	22:23
Air Blank		0.000	22:23
Diagnostics Check		OK	22:24

Data Download Date/Time: 05/06/2019 14:35

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: STACK

DATE/TIME OF TEST: 04/13/2019 02:55:16

Diagnostics Check		OK	02:48
Air Blank		0.000	02:48
Control Test		0.078	02:49
Air Blank		0.000	02:49
Subject Sample	# 1	0.158	02:50
Breath Volume		1.933	02:50
Air Blank		0.000	02:51
Air Blank		0.000	02:53
Subject Sample	# 2	0.158	02:53
Breath Volume		1.964	02:53
Air Blank		0.000	02:54
Control Test		0.078	02:54
Air Blank		0.000	02:55
Diagnostics Check		OK	02:55

Data Download Date/Time: 05/06/2019 14:35

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WEG,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: ELLERBEE

DATE/TIME OF TEST: 04/14/2019 02:51:15

Diagnostics Check		OK	02:43
Air Blank		0.000	02:44
Control Test		0.079	02:44
Air Blank		0.000	02:44
Subject Sample	# 1	0.188	02:46
Breath Volume		1.964	02:46
Air Blank		0.000	02:47
Air Blank		0.000	02:49
Subject Sample	# 2	0.180	02:49
Breath Volume		1.527	02:49
Air Blank		0.000	02:50
Control Test		0.078	02:50
Air Blank		0.000	02:51
Diagnostics Check		OK	02:51

Data Download Date/Time: 05/06/2019 14:35

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: THOMAS,G
ARRESTING OFFICER AGENCY: U of So FL Campus PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: LOUNDERS JR

DATE/TIME OF TEST: 04/14/2019 04:23:07

Diagnostics Check		OK	04:11	
Air Blank		0.000	04:12	
Control Test		0.079	04:12	
Air Blank		0.000	04:13	
Subject Sample	# 1		04:16	No Sample Provided
Breath Volume		0.000	04:16	
Air Blank		0.000	04:16	
Air Blank		0.000	04:18	
Subject Sample	# 2		04:21	No Sample Provided
Breath Volume		0.000	04:21	
Air Blank		0.000	04:22	
Control Test		0.080	04:22	
Air Blank		0.000	04:23	
Diagnostics Check		OK	04:23	

Data Download Date/Time: 05/06/2019 14:35

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CAMPBELL,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:54

SUBJECT'S LAST NAME: KEENEY

DATE/TIME OF TEST: 04/14/2019 22:21:11

Diagnostics Check		OK	22:14
Air Blank		0.000	22:15
Control Test		0.080	22:15
Air Blank		0.000	22:16
Subject Sample	# 1	0.000	22:16
Breath Volume		2.031	22:16
Air Blank		0.000	22:17
Air Blank		0.000	22:19
Subject Sample	# 2	0.000	22:19
Breath Volume		2.007	22:19
Air Blank		0.000	22:20
Control Test		0.079	22:20
Air Blank		0.000	22:21
Diagnostics Check		OK	22:21

Data Download Date/Time: 05/06/2019 14:35

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: THORNE,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: GERVASIO

DATE/TIME OF TEST: 04/15/2019 01:09:33

Diagnostics Check		OK	01:02
Air Blank		0.000	01:03
Control Test		0.080	01:03
Air Blank		0.000	01:03
Subject Sample	# 1	0.000	01:04
Breath Volume		1.257	01:04
Air Blank		0.000	01:04
Air Blank		0.000	01:06
Subject Sample	# 2	0.000	01:08
Breath Volume		1.300	01:08
Air Blank		0.000	01:08
Control Test		0.079	01:08
Air Blank		0.000	01:09
Diagnostics Check		OK	01:09

Data Download Date/Time: 05/06/2019 14:35

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: PEARSON II

DATE/TIME OF TEST: 04/15/2019 03:41:06

Diagnostics Check		OK	03:34
Air Blank		0.000	03:34
Control Test		0.079	03:34
Air Blank		0.000	03:35
Subject Sample	# 1	0.129	03:36
Breath Volume		3.656	03:36
Air Blank		0.000	03:37
Air Blank		0.000	03:38
Subject Sample	# 2	0.129	03:39
Breath Volume		3.730	03:39
Air Blank		0.000	03:40
Control Test		0.079	03:40
Air Blank		0.000	03:41
Diagnostics Check		OK	03:41

Data Download Date/Time: 05/06/2019 14:35

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALL,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 04/16/2019 04:04:54

Diagnostics Check		OK	03:57
Air Blank		0.000	03:57
Control Test		0.079	03:57
Air Blank		0.000	03:58
Subject Sample	# 1	0.157	04:00
Breath Volume		3.015	04:00
Air Blank		0.000	04:00
Air Blank		0.000	04:02
Subject Sample	# 2	0.154	04:03
Breath Volume		2.773	04:03
Air Blank		0.000	04:03
Control Test		0.078	04:04
Air Blank		0.000	04:04
Diagnostics Check		OK	04:04

Data Download Date/Time: 05/06/2019 14:35

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALL,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: KIRNES

DATE/TIME OF TEST: 04/17/2019 01:28:33

Diagnostics Check		OK	01:21
Air Blank		0.000	01:22
Control Test		0.079	01:22
Air Blank		0.000	01:23
Subject Sample	# 1	0.227	01:23
Breath Volume		1.785	01:23
Air Blank		0.000	01:24
Air Blank		0.000	01:26
Subject Sample	# 2	0.214	01:26
Breath Volume		1.535	01:26
Air Blank		0.000	01:27
Control Test		0.079	01:27
Air Blank		0.000	01:28
Diagnostics Check		OK	01:28

Data Download Date/Time: 05/06/2019 14:35

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,ANGELA,V
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MACIAS,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:46

SUBJECT'S LAST NAME: DECKER

DATE/TIME OF TEST: 04/18/2019 18:13:17

Diagnostics Check		OK	18:06
Air Blank		0.000	18:07
Control Test		0.079	18:07
Air Blank		0.000	18:08
Subject Sample	# 1	0.114	18:08
Breath Volume		3.531	18:08
Air Blank		0.000	18:09
Air Blank		0.000	18:11
Subject Sample	# 2	0.108	18:11
Breath Volume		2.691	18:11
Air Blank		0.000	18:12
Control Test		0.079	18:12
Air Blank		0.000	18:13
Diagnostics Check		OK	18:13

Data Download Date/Time: 05/06/2019 14:35

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: WATSON II

DATE/TIME OF TEST: 04/18/2019 23:23:22

Diagnostics Check		OK	23:17
Air Blank		0.000	23:17
Control Test		0.079	23:17
Air Blank		0.000	23:18
Subject Sample	# 1	0.169	23:18
Breath Volume		4.253	23:18
Air Blank		0.000	23:19
Air Blank		0.000	23:21
Subject Sample	# 2	0.170	23:21
Breath Volume		4.824	23:21
Air Blank		0.000	23:22
Control Test		0.078	23:22
Air Blank		0.000	23:23
Diagnostics Check		OK	23:23

Data Download Date/Time: 05/06/2019 14:35

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: NOBLES

DATE/TIME OF TEST: 04/19/2019 01:03:14

Diagnostics Check		OK	00:55
Air Blank		0.000	00:55
Control Test		0.080	00:56
Air Blank		0.000	00:56
Subject Sample	# 1	0.249	00:57
Breath Volume		1.187	00:57
Air Blank		0.000	00:58
Air Blank		0.000	01:00
Subject Sample	# 2	0.264	01:01
Breath Volume		1.886	01:01
Air Blank		0.000	01:02
Control Test		0.078	01:02
Air Blank		0.000	01:03
Diagnostics Check		OK	01:03

Data Download Date/Time: 05/06/2019 14:35

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KLENOFSKY,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: KAMERMAYER

DATE/TIME OF TEST: 04/19/2019 04:18:11

Diagnostics Check		OK	04:10
Air Blank		0.000	04:11
Control Test		0.079	04:11
Air Blank		0.000	04:12
Subject Sample	# 1	0.156	04:13
Breath Volume		2.457	04:13
Air Blank		0.000	04:14
Air Blank		0.000	04:15
Subject Sample	# 2	0.154	04:16
Breath Volume		2.359	04:16
Air Blank		0.000	04:17
Control Test		0.078	04:17
Air Blank		0.000	04:18
Diagnostics Check		OK	04:18

Data Download Date/Time: 05/06/2019 14:35

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LARA,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: REYNEL CASTANON

DATE/TIME OF TEST: 04/19/2019 23:35:52

Diagnostics Check		OK	23:26
Air Blank		0.000	23:27
Control Test		0.079	23:27
Air Blank		0.000	23:28
Subject Sample	# 1	0.188	23:30
Breath Volume		2.140	23:30
Air Blank		0.000	23:31
Air Blank		0.000	23:32
Subject Sample	# 2	0.178	23:34
Breath Volume		1.710	23:34
Air Blank		0.000	23:34
Control Test		0.077	23:35
Air Blank		0.000	23:35
Diagnostics Check		OK	23:35

Data Download Date/Time: 05/06/2019 14:35

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: STAUFFER,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: FIGUEROA

DATE/TIME OF TEST: 04/20/2019 01:15:40

Diagnostics Check		OK	01:07
Air Blank		0.000	01:07
Control Test		0.079	01:08
Air Blank		0.000	01:08
Subject Sample	# 1	0.127	01:09
Breath Volume		1.160	01:09
Air Blank		0.000	01:10
Air Blank		0.000	01:12
Subject Sample	# 2	0.124	01:14
Breath Volume		1.425	01:14
Air Blank		0.000	01:14
Control Test		0.080	01:15
Air Blank		0.000	01:15
Diagnostics Check		OK	01:15

Data Download Date/Time: 05/06/2019 14:35

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DOBSON,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: DUFFEY

DATE/TIME OF TEST: 04/20/2019 01:52:22

Diagnostics Check		OK	01:46
Air Blank		0.000	01:46
Control Test		0.078	01:46
Air Blank		0.000	01:47
Subject Sample	# 1	0.105	01:47
Breath Volume		1.781	01:47
Air Blank		0.000	01:48
Air Blank		0.000	01:50
Subject Sample	# 2	0.103	01:50
Breath Volume		1.660	01:50
Air Blank		0.000	01:51
Control Test		0.078	01:51
Air Blank		0.000	01:52
Diagnostics Check		OK	01:52

Data Download Date/Time: 05/06/2019 14:35

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 04/20/2019 03:48:14

Diagnostics Check		OK	03:41
Air Blank		0.000	03:41
Control Test		0.079	03:42
Air Blank		0.000	03:42
Subject Sample	# 1	0.184	03:43
Breath Volume		3.621	03:43
Air Blank		0.000	03:44
Air Blank		0.000	03:45
Subject Sample	# 2	0.177	03:46
Breath Volume		3.574	03:46
Air Blank		0.000	03:47
Control Test		0.078	03:47
Air Blank		0.000	03:48
Diagnostics Check		OK	03:48

Data Download Date/Time: 05/06/2019 14:35

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES AVERY,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: WARREN

DATE/TIME OF TEST: 04/20/2019 23:15:02

Diagnostics Check		OK	23:09
Air Blank		0.000	23:09
Control Test		0.079	23:09
Air Blank		0.000	23:10
Subject Sample	# 1	0.120	23:10
Breath Volume		2.433	23:10
Air Blank		0.000	23:11
Air Blank		0.000	23:13
Subject Sample	# 2	0.116	23:13
Breath Volume		2.843	23:13
Air Blank		0.000	23:14
Control Test		0.079	23:14
Air Blank		0.000	23:14
Diagnostics Check		OK	23:15

Data Download Date/Time: 05/06/2019 14:35

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CAMPBELL,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: PEREZ-HERNANDEZ

DATE/TIME OF TEST: 04/21/2019 02:26:30

Diagnostics Check		OK	02:18
Air Blank		0.000	02:19
Control Test		0.079	02:19
Air Blank		0.000	02:20
Subject Sample	# 1	0.115	02:22
Breath Volume		1.214	02:22
Air Blank		0.000	02:22
Air Blank		0.000	02:24
Subject Sample	# 2	0.111	02:24
Breath Volume		1.148	02:24
Air Blank		0.000	02:25
Control Test		0.079	02:25
Air Blank		0.000	02:26
Diagnostics Check		OK	02:26

Data Download Date/Time: 05/06/2019 14:35

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES AVERY,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LYON,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: DIAZ CARO

DATE/TIME OF TEST: 04/21/2019 04:40:34

Diagnostics Check		OK	04:33
Air Blank		0.000	04:34
Control Test		0.079	04:34
Air Blank		0.000	04:34
Subject Sample	# 1	0.242	04:36
Breath Volume		1.824	04:36
Air Blank		0.000	04:36
Air Blank		0.000	04:38
Subject Sample	# 2	0.250	04:38
Breath Volume		2.625	04:38
Air Blank		0.000	04:39
Control Test		0.077	04:39
Air Blank		0.000	04:40
Diagnostics Check		OK	04:40

Data Download Date/Time: 05/06/2019 14:35

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES AVERY,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 04/21/2019 23:35:22

Diagnostics Check		OK	23:29
Air Blank		0.000	23:29
Control Test		0.080	23:30
Air Blank		0.000	23:30
Subject Sample	# 1	0.149	23:30
Breath Volume		1.324	23:30
Air Blank		0.000	23:31
Air Blank		0.000	23:33
Subject Sample	# 2	0.163	23:33
Breath Volume		1.261	23:33
Air Blank		0.000	23:34
Control Test		0.079	23:34
Air Blank		0.000	23:35
Diagnostics Check		OK	23:35

Data Download Date/Time: 05/06/2019 14:35

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES AVERY,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALL,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: BOERNER

DATE/TIME OF TEST: 04/22/2019 02:19:08

Diagnostics Check		OK	02:13	
Air Blank		0.000	02:14	
Control Test		0.080	02:14	
Air Blank		0.000	02:15	
Subject Sample	# 1		02:17	Subject Test Refused
Breath Volume		0.000	02:17	
Air Blank		0.000	02:18	
Control Test		0.079	02:18	
Air Blank		0.000	02:19	
Diagnostics Check		OK	02:19	

Data Download Date/Time: 05/06/2019 14:35

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KLENOFSKY,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 04/24/2019 02:46:47

Diagnostics Check		OK	02:39
Air Blank		0.000	02:40
Control Test		0.079	02:40
Air Blank		0.000	02:41
Subject Sample	# 1	0.046	02:42
Breath Volume		2.292	02:42
Air Blank		0.000	02:42
Air Blank		0.000	02:44
Subject Sample	# 2	0.045	02:45
Breath Volume		2.445	02:45
Air Blank		0.000	02:45
Control Test		0.079	02:46
Air Blank		0.000	02:46
Diagnostics Check		OK	02:46

Data Download Date/Time: 05/06/2019 14:35

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 04/25/2019 04:19:18

Diagnostics Check		OK	04:12
Air Blank		0.000	04:13
Control Test		0.078	04:13
Air Blank		0.000	04:13
Subject Sample	# 1	0.163	04:14
Breath Volume		3.234	04:14
Air Blank		0.000	04:15
Air Blank		0.000	04:17
Subject Sample	# 2	0.162	04:17
Breath Volume		3.093	04:17
Air Blank		0.000	04:18
Control Test		0.077	04:18
Air Blank		0.000	04:19
Diagnostics Check		OK	04:19

Data Download Date/Time: 05/06/2019 14:35

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: BELCHER

DATE/TIME OF TEST: 04/25/2019 22:23:33

Diagnostics Check		OK	22:17
Air Blank		0.000	22:17
Control Test		0.079	22:17
Air Blank		0.000	22:18
Subject Sample	# 1	0.219	22:19
Breath Volume		3.062	22:19
Air Blank		0.000	22:19
Air Blank		0.000	22:21
Subject Sample	# 2	0.210	22:21
Breath Volume		2.246	22:21
Air Blank		0.000	22:22
Control Test		0.077	22:22
Air Blank		0.000	22:23
Diagnostics Check		OK	22:23

Data Download Date/Time: 05/06/2019 14:35

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VANTREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: CROSBY II

DATE/TIME OF TEST: 04/26/2019 23:20:23

Diagnostics Check		OK	23:13
Air Blank		0.000	23:14
Control Test		0.078	23:14
Air Blank		0.000	23:15
Subject Sample	# 1	0.214	23:15
Breath Volume		1.531	23:15
Air Blank		0.000	23:16
Air Blank		0.000	23:18
Subject Sample	# 2	0.202	23:18
Breath Volume		1.421	23:18
Air Blank		0.000	23:19
Control Test		0.077	23:19
Air Blank		0.000	23:20
Diagnostics Check		OK	23:20

Data Download Date/Time: 05/06/2019 14:35

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: THORNE,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: SANCHEZ-NAVA

DATE/TIME OF TEST: 04/27/2019 02:45:42

Diagnostics Check		OK	02:33	
Air Blank		0.000	02:34	
Control Test		0.079	02:34	
Air Blank		0.000	02:35	
Subject Sample	# 1	0.165	02:38	Volume Not Met
Breath Volume		0.031	02:38	
Air Blank		0.000	02:39	
Air Blank		0.000	02:40	
Subject Sample	# 2	0.197	02:43	Volume Not Met
Breath Volume		0.175	02:43	
Air Blank		0.000	02:44	
Control Test		0.078	02:45	
Air Blank		0.000	02:45	
Diagnostics Check		OK	02:45	

Data Download Date/Time: 05/06/2019 14:35

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: THORNE,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: SANCHEZ-NAVA

DATE/TIME OF TEST: 04/27/2019 03:06:59

Diagnostics Check		OK	02:53	
Air Blank		0.000	02:53	
Control Test		0.077	02:54	
Air Blank		0.000	02:54	
Subject Sample	# 1	0.196	02:57	Volume Not Met
Breath Volume		0.351	02:57	
Air Blank		0.000	02:58	
Air Blank		0.000	03:00	
Subject Sample	# 2	0.203	03:02	
Breath Volume		1.320	03:02	
Air Blank		0.000	03:02	
Air Blank		0.000	03:04	
Subject Sample	# 3	0.218	03:05	
Breath Volume		1.621	03:05	
Air Blank		0.000	03:06	
Control Test		0.076	03:06	
Air Blank		0.000	03:06	
Diagnostics Check		OK	03:06	

Data Download Date/Time: 05/06/2019 14:35

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: MONTENEGRO

DATE/TIME OF TEST: 04/27/2019 03:17:32

Diagnostics Check		OK	03:11
Air Blank		0.000	03:11
Control Test		0.077	03:11
Air Blank		0.000	03:12
Subject Sample	# 1	0.169	03:13
Breath Volume		2.289	03:13
Air Blank		0.000	03:13
Air Blank		0.000	03:15
Subject Sample	# 2	0.172	03:15
Breath Volume		2.222	03:15
Air Blank		0.000	03:16
Control Test		0.077	03:16
Air Blank		0.000	03:17
Diagnostics Check		OK	03:17

Data Download Date/Time: 05/06/2019 14:35

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: WALZ

DATE/TIME OF TEST: 04/27/2019 21:47:52

Diagnostics Check		OK	21:41
Air Blank		0.000	21:42
Control Test		0.079	21:42
Air Blank		0.000	21:42
Subject Sample	# 1	0.129	21:43
Breath Volume		4.503	21:43
Air Blank		0.000	21:44
Air Blank		0.000	21:45
Subject Sample	# 2	0.130	21:46
Breath Volume		3.203	21:46
Air Blank		0.000	21:46
Control Test		0.078	21:47
Air Blank		0.000	21:47
Diagnostics Check		OK	21:47

Data Download Date/Time: 05/06/2019 14:35

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: VARNUM

DATE/TIME OF TEST: 04/28/2019 02:43:21

Diagnostics Check		OK	02:37
Air Blank		0.000	02:37
Control Test		0.079	02:37
Air Blank		0.000	02:38
Subject Sample	# 1	0.115	02:38
Breath Volume		2.652	02:38
Air Blank		0.000	02:39
Air Blank		0.000	02:41
Subject Sample	# 2	0.114	02:41
Breath Volume		2.304	02:41
Air Blank		0.000	02:42
Control Test		0.079	02:42
Air Blank		0.000	02:43
Diagnostics Check		OK	02:43

Data Download Date/Time: 05/06/2019 14:35

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:51

SUBJECT'S LAST NAME: MUNOZ SIERRA

DATE/TIME OF TEST: 04/28/2019 21:17:46

Diagnostics Check		OK	21:11
Air Blank		0.000	21:11
Control Test		0.079	21:11
Air Blank		0.000	21:12
Subject Sample	# 1	0.180	21:13
Breath Volume		2.753	21:13
Air Blank		0.000	21:13
Air Blank		0.000	21:15
Subject Sample	# 2	0.183	21:16
Breath Volume		3.476	21:16
Air Blank		0.000	21:16
Control Test		0.079	21:17
Air Blank		0.000	21:17
Diagnostics Check		OK	21:17

Data Download Date/Time: 05/06/2019 14:35

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CAMPBELL,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:03

SUBJECT'S LAST NAME: NGUYEN

DATE/TIME OF TEST: 04/29/2019 05:34:54

Diagnostics Check		OK	05:23	
Air Blank		0.000	05:23	
Control Test		0.079	05:23	
Air Blank		0.000	05:24	
Subject Sample	# 1	0.138	05:27	Volume Not Met
Breath Volume		0.394	05:27	
Air Blank		0.000	05:28	
Air Blank		0.000	05:30	
Subject Sample	# 2	0.147	05:33	Volume Not Met
Breath Volume		0.156	05:33	
Air Blank		0.000	05:33	
Control Test		0.078	05:34	
Air Blank		0.000	05:34	
Diagnostics Check		OK	05:34	

Data Download Date/Time: 05/06/2019 14:35

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANDERSON,C
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 04/30/2019 04:38:43

Diagnostics Check		OK	04:29	
Air Blank		0.000	04:30	
Control Test		0.080	04:30	
Air Blank		0.000	04:30	
Subject Sample	# 1		04:34	No Sample Provided
Breath Volume		0.000	04:34	
Air Blank		0.000	04:34	
Air Blank		0.000	04:36	
Subject Sample	# 2		04:37	Subject Test Refused
Breath Volume		0.000	04:37	
Air Blank		0.000	04:37	
Control Test		0.080	04:38	
Air Blank		0.000	04:38	
Diagnostics Check		OK	04:38	

Data Download Date/Time: 05/06/2019 14:35

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: LLOYD

DATE/TIME OF TEST: 04/04/2019 23:33:24

Diagnostics Check		OK	23:27
Air Blank		0.000	23:27
Control Test		0.081	23:27
Air Blank		0.000	23:28
Subject Sample	# 1	0.114	23:28
Breath Volume		1.550	23:28
Air Blank		0.000	23:29
Air Blank		0.000	23:31
Subject Sample	# 2	0.121	23:31
Breath Volume		2.871	23:31
Air Blank		0.000	23:32
Control Test		0.080	23:32
Air Blank		0.000	23:33
Diagnostics Check		OK	23:33

Data Download Date/Time: 05/03/2019 06:23

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CAMPBELL,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: HENRIQUEZ

DATE/TIME OF TEST: 04/05/2019 22:02:04

Diagnostics Check		OK	21:55
Air Blank		0.000	21:56
Control Test		0.081	21:56
Air Blank		0.000	21:56
Subject Sample	# 1	0.081	21:57
Breath Volume		3.796	21:57
Air Blank		0.000	21:58
Air Blank		0.000	22:00
Subject Sample	# 2	0.081	22:00
Breath Volume		4.160	22:00
Air Blank		0.000	22:01
Control Test		0.080	22:01
Air Blank		0.000	22:01
Diagnostics Check		OK	22:02

Data Download Date/Time: 05/03/2019 06:23

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 04/06/2019 05:09:05

Diagnostics Check		OK	05:05	
Air Blank		0.000	05:06	
Control Test		0.080	05:06	
Air Blank		0.000	05:07	
Subject Sample	# 1		05:07	Subject Test Refused
Breath Volume		0.000	05:07	
Air Blank		0.000	05:08	
Control Test		0.080	05:08	
Air Blank		0.000	05:09	
Diagnostics Check		OK	05:09	

Data Download Date/Time: 05/03/2019 06:23

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: VIRTUOSO

DATE/TIME OF TEST: 04/06/2019 22:47:32

Diagnostics Check		OK	22:40
Air Blank		0.000	22:41
Control Test		0.081	22:41
Air Blank		0.000	22:42
Subject Sample	# 1	0.151	22:42
Breath Volume		1.578	22:42
Air Blank		0.000	22:43
Air Blank		0.000	22:45
Subject Sample	# 2	0.160	22:45
Breath Volume		2.660	22:45
Air Blank		0.000	22:46
Control Test		0.080	22:46
Air Blank		0.000	22:47
Diagnostics Check		OK	22:47

Data Download Date/Time: 05/03/2019 06:23

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES AVERY,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SCHAEFER,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: BARRERA SANTANA

DATE/TIME OF TEST: 04/07/2019 01:07:10

Diagnostics Check		OK	01:05	
Air Blank		0.000	01:05	
Control Test		0.080	01:06	
Air Blank		0.000	01:06	
Subject Sample	# 1		01:06	Sequence Aborted
Breath Volume		0.000	01:06	
Air Blank		0.000	01:07	

Data Download Date/Time: 05/03/2019 06:23

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DRISCOLL,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: GARTON III

DATE/TIME OF TEST: 04/07/2019 03:27:27

Diagnostics Check		OK	03:21
Air Blank		0.000	03:21
Control Test		0.080	03:22
Air Blank		0.000	03:22
Subject Sample	# 1	0.009	03:23
Breath Volume		2.996	03:23
Air Blank		0.000	03:23
Air Blank		0.000	03:25
Subject Sample	# 2	0.008	03:26
Breath Volume		3.464	03:26
Air Blank		0.000	03:26
Control Test		0.079	03:26
Air Blank		0.000	03:27
Diagnostics Check		OK	03:27

Data Download Date/Time: 05/03/2019 06:23

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES AVERY,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ARCHAMBEAU,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 04/07/2019 20:39:17

Diagnostics Check		OK	20:32
Air Blank		0.000	20:32
Control Test		0.081	20:33
Air Blank		0.000	20:33
Subject Sample	# 1	0.182	20:34
Breath Volume		1.593	20:34
Air Blank		0.000	20:35
Air Blank		0.000	20:37
Subject Sample	# 2	0.189	20:37
Breath Volume		1.156	20:37
Air Blank		0.000	20:38
Control Test		0.079	20:38
Air Blank		0.000	20:39
Diagnostics Check		OK	20:39

Data Download Date/Time: 05/03/2019 06:23

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANDREWS,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 04/08/2019 00:08:50

Diagnostics Check		OK	23:57	
Air Blank		0.000	23:57	
Control Test		0.080	23:57	
Air Blank		0.000	23:58	
Subject Sample	# 1	0.152	00:01	Volume Not Met
Breath Volume		0.539	00:01	
Air Blank		0.000	00:02	
Air Blank		0.000	00:04	
Subject Sample	# 2	0.150	00:07	Volume Not Met
Breath Volume		0.015	00:07	
Air Blank		0.000	00:07	
Control Test		0.079	00:08	
Air Blank		0.000	00:08	
Diagnostics Check		OK	00:08	

Data Download Date/Time: 05/03/2019 06:23

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANDREWS,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 04/08/2019 00:18:56

Diagnostics Check		OK	00:10	
Air Blank		0.000	00:10	
Control Test		0.079	00:11	
Air Blank		0.000	00:11	
Subject Sample	# 1	0.146	00:14	Volume Not Met
Breath Volume		0.015	00:14	
Air Blank		0.000	00:15	
Air Blank		0.000	00:17	
Subject Sample	# 2		00:17	Subject Test Refused
Breath Volume		0.000	00:17	
Air Blank		0.000	00:18	
Control Test		0.078	00:18	
Air Blank		0.000	00:18	
Diagnostics Check		OK	00:18	

Data Download Date/Time: 05/03/2019 06:23

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MACIAS,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: REYNA-AVILA

DATE/TIME OF TEST: 04/08/2019 22:38:31

Diagnostics Check		OK	22:30
Air Blank		0.000	22:31
Control Test		0.081	22:31
Air Blank		0.000	22:31
Subject Sample	# 1	0.054	22:33
Breath Volume		1.519	22:33
Air Blank		0.000	22:34
Air Blank		0.000	22:36
Subject Sample	# 2	0.059	22:36
Breath Volume		2.820	22:36
Air Blank		0.000	22:37
Control Test		0.081	22:37
Air Blank		0.000	22:38
Diagnostics Check		OK	22:38

Data Download Date/Time: 05/03/2019 06:23

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LYON,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: BRITTICH

DATE/TIME OF TEST: 04/10/2019 00:47:02

Diagnostics Check		OK	00:38
Air Blank		0.000	00:39
Control Test		0.081	00:39
Air Blank		0.000	00:40
Subject Sample	# 1	0.143	00:41
Breath Volume		1.695	00:41
Air Blank		0.000	00:41
Air Blank		0.000	00:43
Subject Sample	# 2	0.144	00:45
Breath Volume		2.195	00:45
Air Blank		0.000	00:46
Control Test		0.078	00:46
Air Blank		0.000	00:46
Diagnostics Check		OK	00:47

Data Download Date/Time: 05/03/2019 06:23

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES AVERY,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: LARRY

DATE/TIME OF TEST: 04/12/2019 01:58:24

Diagnostics Check		OK	01:51
Air Blank		0.000	01:52
Control Test		0.081	01:52
Air Blank		0.000	01:53
Subject Sample	# 1	0.180	01:53
Breath Volume		1.980	01:53
Air Blank		0.000	01:54
Air Blank		0.000	01:56
Subject Sample	# 2	0.182	01:56
Breath Volume		1.875	01:56
Air Blank		0.000	01:57
Control Test		0.079	01:57
Air Blank		0.000	01:58
Diagnostics Check		OK	01:58

Data Download Date/Time: 05/03/2019 06:23

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:17

SUBJECT'S LAST NAME: KREBSBACH

DATE/TIME OF TEST: 04/12/2019 19:45:08

Diagnostics Check		OK	19:37
Air Blank		0.000	19:38
Control Test		0.081	19:38
Air Blank		0.000	19:39
Subject Sample	# 1	0.158	19:40
Breath Volume		1.476	19:40
Air Blank		0.000	19:41
Air Blank		0.000	19:42
Subject Sample	# 2	0.153	19:43
Breath Volume		2.058	19:43
Air Blank		0.000	19:44
Control Test		0.080	19:44
Air Blank		0.000	19:45
Diagnostics Check		OK	19:45

Data Download Date/Time: 05/03/2019 06:23

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KLENOFSKY,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 04/13/2019 04:13:18

Diagnostics Check		OK	04:05
Air Blank		0.000	04:06
Control Test		0.081	04:06
Air Blank		0.000	04:07
Subject Sample	# 1	0.234	04:08
Breath Volume		2.074	04:08
Air Blank		0.000	04:08
Air Blank		0.000	04:10
Subject Sample	# 2	0.221	04:11
Breath Volume		1.238	04:11
Air Blank		0.000	04:12
Control Test		0.079	04:12
Air Blank		0.000	04:13
Diagnostics Check		OK	04:13

Data Download Date/Time: 05/03/2019 06:23

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 06:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:16

SUBJECT'S LAST NAME: FELTENBERGER

DATE/TIME OF TEST: 04/15/2019 06:43:33

Diagnostics Check		OK	06:36
Air Blank		0.000	06:37
Control Test		0.081	06:37
Air Blank		0.000	06:37
Subject Sample	# 1	0.116	06:38
Breath Volume		3.855	06:38
Air Blank		0.000	06:39
Air Blank		0.000	06:41
Subject Sample	# 2	0.117	06:41
Breath Volume		3.167	06:41
Air Blank		0.000	06:42
Control Test		0.080	06:42
Air Blank		0.000	06:43
Diagnostics Check		OK	06:43

Data Download Date/Time: 05/03/2019 06:23

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SCHAEFER,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: LEON ARANDA

DATE/TIME OF TEST: 04/16/2019 03:54:44

Diagnostics Check		OK	03:47
Air Blank		0.000	03:48
Control Test		0.081	03:48
Air Blank		0.000	03:48
Subject Sample	# 1	0.198	03:50
Breath Volume		2.300	03:50
Air Blank		0.000	03:50
Air Blank		0.000	03:52
Subject Sample	# 2	0.199	03:53
Breath Volume		1.410	03:53
Air Blank		0.000	03:53
Control Test		0.079	03:54
Air Blank		0.000	03:54
Diagnostics Check		OK	03:54

Data Download Date/Time: 05/03/2019 06:23

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES AVERY,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WEG,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: DOMINGUEZ RIOS

DATE/TIME OF TEST: 04/18/2019 00:23:12

Diagnostics Check		OK	00:17
Air Blank		0.000	00:17
Control Test		0.080	00:17
Air Blank		0.000	00:18
Subject Sample	# 1	0.149	00:18
Breath Volume		3.046	00:18
Air Blank		0.000	00:19
Air Blank		0.000	00:21
Subject Sample	# 2	0.139	00:21
Breath Volume		1.933	00:21
Air Blank		0.000	00:22
Control Test		0.079	00:22
Air Blank		0.000	00:23
Diagnostics Check		OK	00:23

Data Download Date/Time: 05/03/2019 06:23

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CAMPBELL,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: TRIPLETT

DATE/TIME OF TEST: 04/19/2019 05:01:15

Diagnostics Check		OK	04:54
Air Blank		0.000	04:54
Control Test		0.081	04:54
Air Blank		0.000	04:55
Subject Sample	# 1	0.157	04:56
Breath Volume		1.371	04:56
Air Blank		0.000	04:56
Air Blank		0.000	04:58
Subject Sample	# 2	0.154	04:59
Breath Volume		2.097	04:59
Air Blank		0.000	05:00
Control Test		0.080	05:00
Air Blank		0.000	05:01
Diagnostics Check		OK	05:01

Data Download Date/Time: 05/03/2019 06:23

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,ANGELA,V
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANGELICCH,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:28

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 04/19/2019 17:56:08

Diagnostics Check		OK	17:49
Air Blank		0.000	17:50
Control Test		0.080	17:50
Air Blank		0.000	17:51
Subject Sample	# 1	0.238	17:51
Breath Volume		2.226	17:51
Air Blank		0.000	17:52
Air Blank		0.000	17:54
Subject Sample	# 2	0.242	17:54
Breath Volume		2.359	17:54
Air Blank		0.000	17:55
Control Test		0.079	17:55
Air Blank		0.000	17:56
Diagnostics Check		OK	17:56

Data Download Date/Time: 05/03/2019 06:23

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: PADILLA

DATE/TIME OF TEST: 04/19/2019 23:08:06

Diagnostics Check		OK	23:01
Air Blank		0.000	23:01
Control Test		0.081	23:01
Air Blank		0.000	23:02
Subject Sample	# 1	0.146	23:03
Breath Volume		2.019	23:03
Air Blank		0.000	23:04
Air Blank		0.000	23:06
Subject Sample	# 2	0.136	23:06
Breath Volume		1.187	23:06
Air Blank		0.000	23:07
Control Test		0.080	23:07
Air Blank		0.000	23:08
Diagnostics Check		OK	23:08

Data Download Date/Time: 05/03/2019 06:23

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: LINDE

DATE/TIME OF TEST: 04/20/2019 02:37:30

Diagnostics Check		OK	02:24	
Air Blank		0.000	02:24	
Control Test		0.081	02:24	
Air Blank		0.000	02:25	
Subject Sample	# 1	0.108	02:28	Volume Not Met
Breath Volume		0.058	02:28	
Air Blank		0.000	02:29	
Air Blank		0.000	02:30	
Subject Sample	# 2	0.138	02:32	
Breath Volume		2.894	02:32	
Air Blank		0.000	02:33	
Air Blank		0.000	02:35	
Subject Sample	# 3	0.131	02:35	
Breath Volume		1.703	02:35	
Air Blank		0.000	02:36	
Control Test		0.077	02:36	
Air Blank		0.000	02:37	
Diagnostics Check		OK	02:37	

Data Download Date/Time: 05/03/2019 06:23

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VANTREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: GRIFFIN JR

DATE/TIME OF TEST: 04/20/2019 04:03:10

Diagnostics Check		OK	03:50	
Air Blank		0.000	03:51	
Control Test		0.080	03:51	
Air Blank		0.000	03:51	
Subject Sample	# 1	0.158	03:55	Volume Not Met
Breath Volume		0.531	03:55	
Air Blank		0.000	03:55	
Air Blank		0.000	03:57	
Subject Sample	# 2	0.164	03:58	
Breath Volume		1.867	03:58	
Air Blank		0.000	03:59	
Air Blank		0.000	04:00	
Subject Sample	# 3	0.163	04:01	
Breath Volume		1.621	04:01	
Air Blank		0.000	04:02	
Control Test		0.078	04:02	
Air Blank		0.000	04:03	
Diagnostics Check		OK	04:03	

Data Download Date/Time: 05/03/2019 06:24

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VASQUEZ,J
ARRESTING OFFICER AGENCY: Tampa Airport PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:56

SUBJECT'S LAST NAME: CORONADO GONZALEZ

DATE/TIME OF TEST: 04/20/2019 05:24:36

Diagnostics Check		OK	05:16
Air Blank		0.000	05:17
Control Test		0.080	05:17
Air Blank		0.000	05:18
Subject Sample	# 1	0.162	05:19
Breath Volume		1.121	05:19
Air Blank		0.000	05:20
Air Blank		0.000	05:22
Subject Sample	# 2	0.174	05:23
Breath Volume		2.082	05:23
Air Blank		0.000	05:23
Control Test		0.079	05:23
Air Blank		0.000	05:24
Diagnostics Check		OK	05:24

Data Download Date/Time: 05/03/2019 06:24

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SCHAEFER,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: PEREZ-HERNANDEZ

DATE/TIME OF TEST: 04/21/2019 01:34:26

Diagnostics Check		OK	01:27
Air Blank		0.000	01:28
Control Test		0.081	01:28
Air Blank		0.000	01:29
Subject Sample	# 1	0.202	01:29
Breath Volume		2.070	01:29
Air Blank		0.000	01:30
Air Blank		0.000	01:32
Subject Sample	# 2	0.203	01:32
Breath Volume		2.218	01:32
Air Blank		0.000	01:33
Control Test		0.080	01:33
Air Blank		0.000	01:34
Diagnostics Check		OK	01:34

Data Download Date/Time: 05/03/2019 06:24

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES AVERY,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 04/21/2019 03:57:30

Diagnostics Check		OK	03:51
Air Blank		0.000	03:51
Control Test		0.080	03:52
Air Blank		0.000	03:52
Subject Sample	# 1	0.107	03:53
Breath Volume		2.910	03:53
Air Blank		0.000	03:53
Air Blank		0.000	03:55
Subject Sample	# 2	0.100	03:55
Breath Volume		2.082	03:55
Air Blank		0.000	03:56
Control Test		0.079	03:56
Air Blank		0.000	03:57
Diagnostics Check		OK	03:57

Data Download Date/Time: 05/03/2019 06:24

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES AVERY,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DOBSON,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: PALOMO

DATE/TIME OF TEST: 04/21/2019 23:45:39

Diagnostics Check		OK	23:39
Air Blank		0.000	23:39
Control Test		0.080	23:40
Air Blank		0.000	23:40
Subject Sample	# 1	0.148	23:41
Breath Volume		1.402	23:41
Air Blank		0.000	23:41
Air Blank		0.000	23:43
Subject Sample	# 2	0.154	23:44
Breath Volume		1.261	23:44
Air Blank		0.000	23:44
Control Test		0.082	23:45
Air Blank		0.000	23:45
Diagnostics Check		OK	23:45

Data Download Date/Time: 05/03/2019 06:24

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: THORNE,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 04/23/2019 05:09:58

Diagnostics Check		OK	05:03
Air Blank		0.000	05:04
Control Test		0.082	05:04
Air Blank		0.000	05:04
Subject Sample	# 1	0.028	05:05
Breath Volume		1.156	05:05
Air Blank		0.000	05:06
Air Blank		0.000	05:07
Subject Sample	# 2	0.028	05:08
Breath Volume		1.300	05:08
Air Blank		0.000	05:09
Control Test		0.080	05:09
Air Blank		0.000	05:09
Diagnostics Check		OK	05:09

Data Download Date/Time: 05/03/2019 06:24

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: ABUDAYAH

DATE/TIME OF TEST: 04/25/2019 02:26:03

Diagnostics Check		OK	02:19
Air Blank		0.000	02:20
Control Test		0.081	02:20
Air Blank		0.000	02:21
Subject Sample	# 1	0.000	02:21
Breath Volume		2.589	02:21
Air Blank		0.000	02:22
Air Blank		0.000	02:24
Subject Sample	# 2	0.000	02:24
Breath Volume		3.148	02:24
Air Blank		0.000	02:25
Control Test		0.081	02:25
Air Blank		0.000	02:25
Diagnostics Check		OK	02:26

Data Download Date/Time: 05/03/2019 06:24

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WHITE,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: MROCZEK

DATE/TIME OF TEST: 04/26/2019 02:56:38

Diagnostics Check		OK	02:50
Air Blank		0.000	02:50
Control Test		0.081	02:50
Air Blank		0.000	02:51
Subject Sample	# 1	0.210	02:51
Breath Volume		3.574	02:51
Air Blank		0.000	02:52
Air Blank		0.000	02:54
Subject Sample	# 2	0.193	02:55
Breath Volume		1.472	02:55
Air Blank		0.000	02:55
Control Test		0.080	02:56
Air Blank		0.000	02:56
Diagnostics Check		OK	02:56

Data Download Date/Time: 05/03/2019 06:24

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 04/26/2019 21:21:21

Diagnostics Check		OK	21:13
Air Blank		0.000	21:14
Control Test		0.080	21:14
Air Blank		0.000	21:15
Subject Sample	# 1	0.154	21:16
Breath Volume		1.394	21:16
Air Blank		0.000	21:17
Air Blank		0.000	21:19
Subject Sample	# 2	0.148	21:19
Breath Volume		1.617	21:19
Air Blank		0.000	21:20
Control Test		0.079	21:20
Air Blank		0.000	21:21
Diagnostics Check		OK	21:21

Data Download Date/Time: 05/03/2019 06:24

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: KANHAI

DATE/TIME OF TEST: 04/27/2019 04:23:50

Diagnostics Check		OK	04:17
Air Blank		0.000	04:18
Control Test		0.080	04:18
Air Blank		0.000	04:18
Subject Sample	# 1	0.148	04:19
Breath Volume		1.164	04:19
Air Blank		0.000	04:19
Air Blank		0.000	04:21
Subject Sample	# 2	0.145	04:22
Breath Volume		2.039	04:22
Air Blank		0.000	04:22
Control Test		0.080	04:23
Air Blank		0.000	04:23
Diagnostics Check		OK	04:23

Data Download Date/Time: 05/03/2019 06:24

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LYON,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 04/27/2019 22:57:58

Diagnostics Check		OK	22:48
Air Blank		0.000	22:48
Control Test		0.081	22:49
Air Blank		0.000	22:49
Subject Sample	# 1	0.235	22:52
Breath Volume		1.132	22:52
Air Blank		0.000	22:52
Air Blank		0.000	22:54
Subject Sample	# 2	0.233	22:56
Breath Volume		1.574	22:56
Air Blank		0.000	22:57
Control Test		0.077	22:57
Air Blank		0.000	22:57
Diagnostics Check		OK	22:57

Data Download Date/Time: 05/03/2019 06:24

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KLENOFSKY,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: GARCIA ONTIVEROS

DATE/TIME OF TEST: 04/28/2019 03:15:55

Diagnostics Check		OK	03:07
Air Blank		0.000	03:07
Control Test		0.081	03:08
Air Blank		0.000	03:08
Subject Sample	# 1	0.322	03:09
Breath Volume		1.226	03:09
Air Blank		0.000	03:10
Air Blank		0.000	03:11
Subject Sample	# 2	0.323	03:14
Breath Volume		1.261	03:14
Air Blank		0.000	03:14
Control Test		0.080	03:15
Air Blank		0.000	03:15
Diagnostics Check		OK	03:15

Data Download Date/Time: 05/03/2019 06:24

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DRISCOLL,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:17

SUBJECT'S LAST NAME: TRAY

DATE/TIME OF TEST: 04/29/2019 21:49:41

Diagnostics Check		OK	21:37	
Air Blank		0.000	21:38	
Control Test		0.081	21:38	
Air Blank		0.000	21:39	
Subject Sample	# 1	0.158	21:42	Volume Not Met
Breath Volume		0.050	21:42	
Air Blank		0.000	21:43	
Air Blank		0.000	21:44	
Subject Sample	# 2	0.146	21:48	Volume Not Met
Breath Volume		0.265	21:48	
Air Blank		0.000	21:48	
Control Test		0.078	21:49	
Air Blank		0.000	21:49	
Diagnostics Check		OK	21:49	

Data Download Date/Time: 05/03/2019 06:24

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DRISCOLL,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:17

SUBJECT'S LAST NAME: TRAY

DATE/TIME OF TEST: 04/29/2019 22:00:10

Diagnostics Check		OK	21:52
Air Blank		0.000	21:52
Control Test		0.079	21:52
Air Blank		0.000	21:53
Subject Sample	# 1	0.175	21:54
Breath Volume		2.437	21:54
Air Blank		0.000	21:55
Air Blank		0.000	21:57
Subject Sample	# 2	0.160	21:58
Breath Volume		1.726	21:58
Air Blank		0.000	21:59
Control Test		0.077	21:59
Air Blank		0.000	22:00
Diagnostics Check		OK	22:00

Data Download Date/Time: 05/03/2019 06:24

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003407

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CABRAL,JONATHAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: ERNST,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:46

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 04/14/2019 21:18:44

Diagnostics Check		OK	21:11
Air Blank		0.000	21:12
Control Test		0.078	21:12
Air Blank		0.000	21:13
Subject Sample	# 1	0.158	21:14
Breath Volume		1.988	21:14
Air Blank		0.000	21:14
Air Blank		0.000	21:16
Subject Sample	# 2	0.155	21:17
Breath Volume		1.445	21:17
Air Blank		0.000	21:17
Control Test		0.077	21:18
Air Blank		0.000	21:18
Diagnostics Check		OK	21:18

Data Download Date/Time: 04/27/2019 23:53

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ-CORTINA,RAUL,H
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: GONZALEZ-CORTINA,RAUL
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 620051
DRY GAS EXPIRY DATE: 11/05/2019
VIOLATION CODE: DUI
ARREST TIME: 15:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:53

SUBJECT'S LAST NAME: NICOLAS

DATE/TIME OF TEST: 04/05/2019 16:19:36

Diagnostics Check		OK	16:13
Air Blank		0.000	16:14
Control Test		0.080	16:14
Air Blank		0.000	16:14
Subject Sample	# 1	0.187	16:15
Breath Volume		3.382	16:15
Air Blank		0.000	16:15
Air Blank		0.000	16:17
Subject Sample	# 2	0.183	16:18
Breath Volume		3.746	16:18
Air Blank		0.000	16:18
Control Test		0.080	16:18
Air Blank		0.000	16:19
Diagnostics Check		OK	16:19

Data Download Date/Time: 05/01/2019 00:53

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LA ROSA CASTRO,DAVID,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: FERNANDEZ,ORLANDO
ARRESTING OFFICER AGENCY: SWEETWATER PD
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 620051
DRY GAS EXPIRY DATE: 11/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:42

SUBJECT'S LAST NAME: PEINADO

DATE/TIME OF TEST: 04/06/2019 21:06:14

Diagnostics Check		OK	21:03	
Air Blank		0.000	21:03	
Control Test		0.080	21:04	
Air Blank		0.000	21:04	
Subject Sample	# 1		21:04	Subject Test Refused
Breath Volume		0.000	21:04	
Air Blank		0.000	21:05	
Control Test		0.080	21:05	
Air Blank		0.000	21:06	
Diagnostics Check		OK	21:06	

Data Download Date/Time: 05/01/2019 00:53

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANCHEZ,JUSTIN,A
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: SANCHEZ,JUSTIN
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 620051
DRY GAS EXPIRY DATE: 11/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: FACEY

DATE/TIME OF TEST: 04/10/2019 03:53:24

Diagnostics Check	OK	03:52	
Air Blank		03:52	Improper Sample
Air Blank	0.000	03:53	

Data Download Date/Time: 05/01/2019 00:53

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANCHEZ,JUSTIN,A
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: SANCHEZ,JUSTIN
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 620051
DRY GAS EXPIRY DATE: 11/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: FACEY

DATE/TIME OF TEST: 04/10/2019 04:05:11

Diagnostics Check		OK	03:57
Air Blank		0.000	03:57
Control Test		0.080	03:57
Air Blank		0.000	03:58
Subject Sample	# 1	0.133	04:00
Breath Volume		1.214	04:00
Air Blank		0.000	04:01
Air Blank		0.000	04:03
Subject Sample	# 2	0.136	04:03
Breath Volume		1.281	04:03
Air Blank		0.000	04:04
Control Test		0.077	04:04
Air Blank		0.000	04:05
Diagnostics Check		OK	04:05

Data Download Date/Time: 05/01/2019 00:53

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TSOI,YOQUEL,E
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: TSOI,YOQUEL
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 620051
DRY GAS EXPIRY DATE: 11/05/2019
VIOLATION CODE: DUI
ARREST TIME: 08:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 04/11/2019 22:06:19

Diagnostics Check		OK	21:53	
Air Blank		0.000	21:54	
Control Test		0.081	21:54	
Air Blank		0.000	21:54	
Subject Sample	# 1	0.298	21:58	Volume Not Met
Breath Volume		0.519	21:58	
Air Blank		0.000	21:58	
Air Blank		0.000	22:00	
Subject Sample	# 2	0.286	22:01	
Breath Volume		1.179	22:01	
Air Blank		0.000	22:01	
Air Blank		0.000	22:03	
Subject Sample	# 3	0.295	22:04	
Breath Volume		1.238	22:04	
Air Blank		0.000	22:05	
Control Test		0.078	22:05	
Air Blank		0.000	22:06	
Diagnostics Check		OK	22:06	

Data Download Date/Time: 05/01/2019 00:53

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JUDAH,DAVID,B
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: JUDAH,DAVID
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 620051
DRY GAS EXPIRY DATE: 11/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: FEREIRA

DATE/TIME OF TEST: 04/14/2019 03:18:43

Diagnostics Check		OK	03:12
Air Blank		0.000	03:12
Control Test		0.081	03:13
Air Blank		0.000	03:13
Subject Sample	# 1	0.063	03:14
Breath Volume		2.910	03:14
Air Blank		0.000	03:14
Air Blank		0.000	03:16
Subject Sample	# 2	0.058	03:17
Breath Volume		1.894	03:17
Air Blank		0.000	03:17
Control Test		0.081	03:18
Air Blank		0.000	03:18
Diagnostics Check		OK	03:18

Data Download Date/Time: 05/01/2019 00:53

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPOTE,DUVAN,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: CAPOTE,DUVAN
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 620051
DRY GAS EXPIRY DATE: 11/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: MANA

DATE/TIME OF TEST: 04/14/2019 19:01:48

Diagnostics Check		OK	18:55
Air Blank		0.000	18:56
Control Test		0.080	18:56
Air Blank		0.000	18:57
Subject Sample	# 1	0.000	18:57
Breath Volume		2.031	18:57
Air Blank		0.000	18:58
Air Blank		0.000	19:00
Subject Sample	# 2	0.000	19:00
Breath Volume		2.464	19:00
Air Blank		0.000	19:00
Control Test		0.079	19:01
Air Blank		0.000	19:01
Diagnostics Check		OK	19:01

Data Download Date/Time: 05/01/2019 00:53

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JUDAH,DAVID,B
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: JUDAH,DAVID
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 620051
DRY GAS EXPIRY DATE: 11/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: QUITIAN

DATE/TIME OF TEST: 04/18/2019 00:47:56

Diagnostics Check		OK	00:41
Air Blank		0.000	00:42
Control Test		0.081	00:42
Air Blank		0.000	00:43
Subject Sample	# 1	0.104	00:43
Breath Volume		2.621	00:43
Air Blank		0.000	00:44
Air Blank		0.000	00:46
Subject Sample	# 2	0.119	00:46
Breath Volume		4.289	00:46
Air Blank		0.000	00:47
Control Test		0.080	00:47
Air Blank		0.000	00:47
Diagnostics Check		OK	00:47

Data Download Date/Time: 05/01/2019 00:53

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,ARGENIS,A
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: PEREZ,ARGENIS
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 620051
DRY GAS EXPIRY DATE: 11/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: GUTIERREZ

DATE/TIME OF TEST: 04/21/2019 23:22:03

Diagnostics Check		OK	23:19	
Air Blank		0.000	23:19	
Control Test		0.080	23:19	
Air Blank		0.000	23:20	
Subject Sample	# 1		23:20	Subject Test Refused
Breath Volume		0.000	23:20	
Air Blank		0.000	23:21	
Control Test		0.081	23:21	
Air Blank		0.000	23:21	
Diagnostics Check		OK	23:22	

Data Download Date/Time: 05/01/2019 00:53

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NANAN,KEIL,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: NANAN,KEIL
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 620051
DRY GAS EXPIRY DATE: 11/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:00

SUBJECT'S LAST NAME: BECKER TAMAYO

DATE/TIME OF TEST: 04/22/2019 07:23:43

Diagnostics Check		OK	07:20	
Air Blank		0.000	07:21	
Control Test		0.080	07:21	
Air Blank		0.000	07:22	
Subject Sample	# 1		07:22	Subject Test Refused
Breath Volume		0.000	07:22	
Air Blank		0.000	07:22	
Control Test		0.081	07:23	
Air Blank		0.000	07:23	
Diagnostics Check		OK	07:23	

Data Download Date/Time: 05/01/2019 00:53

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALBO,JOSHUA,D
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: GIZ,JOAQUIN
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 620051
DRY GAS EXPIRY DATE: 11/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: K126000724520

DATE/TIME OF TEST: 04/24/2019 20:29:05

Diagnostics Check		OK	20:25	
Air Blank		0.000	20:26	
Control Test		0.081	20:26	
Air Blank		0.000	20:26	
Subject Sample	# 1		20:27	Subject Test Refused
Breath Volume		0.000	20:27	
Air Blank		0.000	20:28	
Control Test		0.080	20:28	
Air Blank		0.000	20:29	
Diagnostics Check		OK	20:29	

Data Download Date/Time: 05/01/2019 00:53

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTILLO,MIDALYS,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: ESTEVEZ,JOSE
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 620051
DRY GAS EXPIRY DATE: 11/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 04/25/2019 22:32:13

Diagnostics Check		OK	22:29	
Air Blank		0.000	22:29	
Control Test		0.080	22:30	
Air Blank		0.000	22:30	
Subject Sample	# 1		22:30	Subject Test Refused
Breath Volume		0.000	22:30	
Air Blank		0.000	22:31	
Control Test		0.081	22:31	
Air Blank		0.000	22:32	
Diagnostics Check		OK	22:32	

Data Download Date/Time: 05/01/2019 00:54

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESTEVEZ,JOSE,M
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: ESTEVEZ,JOSE
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 620051
DRY GAS EXPIRY DATE: 11/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 04/25/2019 22:38:38

Diagnostics Check		OK	22:35	
Air Blank		0.000	22:36	
Control Test		0.081	22:36	
Air Blank		0.000	22:37	
Subject Sample	# 1		22:37	Subject Test Refused
Breath Volume		0.000	22:37	
Air Blank		0.000	22:37	
Control Test		0.080	22:38	
Air Blank		0.000	22:38	
Diagnostics Check		OK	22:38	

Data Download Date/Time: 05/01/2019 00:54

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSQUERA,ODEL,J
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: MOSQUERA,ODEL
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 620051
DRY GAS EXPIRY DATE: 11/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: GUMBS

DATE/TIME OF TEST: 04/30/2019 02:50:51

Diagnostics Check		OK	02:47	
Air Blank		0.000	02:48	
Control Test		0.081	02:48	
Air Blank		0.000	02:49	
Subject Sample	# 1		02:49	Subject Test Refused
Breath Volume		0.000	02:49	
Air Blank		0.000	02:49	
Control Test		0.080	02:50	
Air Blank		0.000	02:50	
Diagnostics Check		OK	02:50	

Data Download Date/Time: 05/01/2019 00:54

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003410

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESTELUS,JUNIOR,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: ESTELUS,JUNIOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 437738
DRY GAS EXPIRY DATE: 06/08/2019
VIOLATION CODE: DUI
ARREST TIME: 19:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: BRUNS

DATE/TIME OF TEST: 04/19/2019 21:01:22

Diagnostics Check	OK	21:00	
Air Blank	0.000	21:01	
		21:01	RFI Detect
Air Blank		21:01	RFI Detect

Data Download Date/Time: 04/22/2019 17:54

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003410

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESTELUS,JUNIOR,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: ESTELUS,JUNIOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 437738
DRY GAS EXPIRY DATE: 06/08/2019
VIOLATION CODE: DUI
ARREST TIME: 19:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: BRUNS

DATE/TIME OF TEST: 04/19/2019 21:06:34

Diagnostics Check	Fail	21:06	Voltage/Current Test
Air Blank	0.000	21:06	

Data Download Date/Time: 04/22/2019 17:54

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003410

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESTELUS,JUNIOR,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: ESTELUS,JUNIOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 437738
DRY GAS EXPIRY DATE: 06/08/2019
VIOLATION CODE: DUI
ARREST TIME: 19:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: BRUNS

DATE/TIME OF TEST: 04/19/2019 21:16:21

Diagnostics Check		OK	21:10
Air Blank		0.000	21:10
Control Test		0.079	21:10
Air Blank		0.000	21:11
Subject Sample	# 1	0.121	21:12
Breath Volume		3.160	21:12
Air Blank		0.000	21:12
Air Blank		0.000	21:14
Subject Sample	# 2	0.117	21:14
Breath Volume		4.464	21:14
Air Blank		0.000	21:15
Control Test		0.081	21:15
Air Blank		0.000	21:16
Diagnostics Check		OK	21:16

Data Download Date/Time: 04/22/2019 17:54

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003412

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: STRAHAN

DATE/TIME OF TEST: 04/05/2019 21:36:48

Diagnostics Check		OK	21:30
Air Blank		0.000	21:30
Control Test		0.081	21:31
Air Blank		0.000	21:31
Subject Sample	# 1	0.093	21:32
Breath Volume		1.148	21:32
Air Blank		0.000	21:32
Air Blank		0.000	21:34
Subject Sample	# 2	0.102	21:35
Breath Volume		2.597	21:35
Air Blank		0.000	21:35
Control Test		0.081	21:36
Air Blank		0.000	21:36
Diagnostics Check		OK	21:36

Data Download Date/Time: 04/24/2019 18:13

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003412

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 19:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: FRANCE

DATE/TIME OF TEST: 04/06/2019 20:16:27

Diagnostics Check		OK	20:10
Air Blank		0.000	20:10
Control Test		0.080	20:11
Air Blank		0.000	20:11
Subject Sample	# 1	0.000	20:12
Breath Volume		3.093	20:12
Air Blank		0.000	20:12
Air Blank		0.000	20:14
Subject Sample	# 2	0.000	20:15
Breath Volume		2.957	20:15
Air Blank		0.000	20:15
Control Test		0.080	20:15
Air Blank		0.000	20:16
Diagnostics Check		OK	20:16

Data Download Date/Time: 04/24/2019 18:13

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003412

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: BOZIK

DATE/TIME OF TEST: 04/08/2019 00:29:38

Diagnostics Check		OK	00:23
Air Blank		0.000	00:23
Control Test		0.080	00:23
Air Blank		0.000	00:24
Subject Sample	# 1	0.145	00:24
Breath Volume		2.761	00:24
Air Blank		0.000	00:25
Air Blank		0.000	00:27
Subject Sample	# 2	0.139	00:28
Breath Volume		2.207	00:28
Air Blank		0.000	00:28
Control Test		0.080	00:29
Air Blank		0.000	00:29
Diagnostics Check		OK	00:29

Data Download Date/Time: 04/24/2019 18:13

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003412

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,A
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 03/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: FERNANDES

DATE/TIME OF TEST: 04/08/2019 02:40:05

Diagnostics Check	OK	02:38	
Air Blank	0.000	02:39	
Control Test	0.080	02:39	Interferent Detect
Air Blank	0.000	02:40	

Data Download Date/Time: 04/24/2019 18:13

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003412

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,A,M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,A
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 03/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: FERNANDES

DATE/TIME OF TEST: 04/08/2019 02:43:35

Diagnostics Check	OK	02:42	
Air Blank	0.000	02:42	
Control Test	0.080	02:43	Interferent Detect
Air Blank	0.000	02:43	

Data Download Date/Time: 04/24/2019 18:13

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003412

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: BARNES

DATE/TIME OF TEST: 04/11/2019 23:36:35

Diagnostics Check		OK	23:30
Air Blank		0.000	23:30
Control Test		0.080	23:31
Air Blank		0.000	23:31
Subject Sample	# 1	0.121	23:32
Breath Volume		3.019	23:32
Air Blank		0.000	23:32
Air Blank		0.000	23:34
Subject Sample	# 2	0.120	23:35
Breath Volume		2.878	23:35
Air Blank		0.000	23:35
Control Test		0.081	23:35
Air Blank		0.000	23:36
Diagnostics Check		OK	23:36

Data Download Date/Time: 04/24/2019 18:13

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003412

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: TRUSCIO,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 03/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: REESE

DATE/TIME OF TEST: 04/15/2019 23:51:55

Diagnostics Check		OK	23:45
Air Blank		0.000	23:46
Control Test		0.081	23:46
Air Blank		0.000	23:46
Subject Sample	# 1	0.210	23:47
Breath Volume		1.320	23:47
Air Blank		0.000	23:48
Air Blank		0.000	23:49
Subject Sample	# 2	0.211	23:50
Breath Volume		3.773	23:50
Air Blank		0.000	23:50
Control Test		0.081	23:51
Air Blank		0.000	23:51
Diagnostics Check		OK	23:51

Data Download Date/Time: 04/24/2019 18:13

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JARRETT,RASHAHN,C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JARRETT,RASHAHN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 11:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:20

SUBJECT'S LAST NAME: COON

DATE/TIME OF TEST: 04/29/2019 11:47:54

Diagnostics Check		OK	11:40
Air Blank		0.000	11:41
Control Test		0.080	11:41
Air Blank		0.000	11:42
Subject Sample	# 1	0.239	11:43
Breath Volume		1.855	11:43
Air Blank		0.000	11:44
Air Blank		0.000	11:45
Subject Sample	# 2	0.230	11:46
Breath Volume		1.359	11:46
Air Blank		0.000	11:47
Control Test		0.079	11:47
Air Blank		0.000	11:47
Diagnostics Check		OK	11:47

Data Download Date/Time: 05/23/2019 14:24

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PYE,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: PYE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 04/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: DE BERGH-IRVING

DATE/TIME OF TEST: 04/05/2019 04:12:26

Diagnostics Check		OK	04:06
Air Blank		0.000	04:06
Control Test		0.083	04:07
Air Blank		0.000	04:07
Subject Sample	# 1	0.025	04:08
Breath Volume		1.671	04:08
Air Blank		0.000	04:08
Air Blank		0.000	04:10
Subject Sample	# 2	0.024	04:10
Breath Volume		1.421	04:10
Air Blank		0.000	04:11
Control Test		0.081	04:11
Air Blank		0.000	04:12
Diagnostics Check		OK	04:12

Data Download Date/Time: 05/02/2019 05:12

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAYERS,MICHELLE,H
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: SAYERS,MICHELLE
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 04/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: MALDONADO-CORTES

DATE/TIME OF TEST: 04/12/2019 00:19:45

Diagnostics Check		OK	00:12
Air Blank		0.000	00:13
Control Test		0.081	00:13
Air Blank		0.000	00:14
Subject Sample	# 1	0.308	00:15
Breath Volume		3.242	00:15
Air Blank		0.000	00:15
Air Blank		0.000	00:17
Subject Sample	# 2	0.308	00:18
Breath Volume		3.136	00:18
Air Blank		0.000	00:18
Control Test		0.081	00:19
Air Blank		0.000	00:19
Diagnostics Check		OK	00:19

Data Download Date/Time: 05/02/2019 05:12

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAYERS,MICHELLE,H
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: SAYERS,MICHELLE
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 04/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: IRIZARRY

DATE/TIME OF TEST: 04/20/2019 00:07:01

Diagnostics Check		OK	00:00
Air Blank		0.000	00:00
Control Test		0.082	00:00
Air Blank		0.000	00:01
Subject Sample	# 1	0.144	00:02
Breath Volume		3.957	00:02
Air Blank		0.000	00:02
Air Blank		0.000	00:04
Subject Sample	# 2	0.149	00:05
Breath Volume		3.027	00:05
Air Blank		0.000	00:06
Control Test		0.083	00:06
Air Blank		0.000	00:06
Diagnostics Check		OK	00:07

Data Download Date/Time: 05/02/2019 05:12

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAYERS,MICHELLE,H
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: SAYERS,MICHELLE
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 04/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: JEROME

DATE/TIME OF TEST: 04/30/2019 01:51:52

Diagnostics Check		OK	01:45
Air Blank		0.000	01:45
Control Test		0.081	01:46
Air Blank		0.000	01:46
Subject Sample	# 1	0.000	01:47
Breath Volume		2.671	01:47
Air Blank		0.000	01:48
Air Blank		0.000	01:50
Subject Sample	# 2	0.000	01:50
Breath Volume		3.722	01:50
Air Blank		0.000	01:50
Control Test		0.081	01:51
Air Blank		0.000	01:51
Diagnostics Check		OK	01:51

Data Download Date/Time: 05/02/2019 05:12

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAYERS,MICHELLE,H
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: SAYERS,MICHELLE
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 04/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: PESANTE-ORTIZ

DATE/TIME OF TEST: 04/30/2019 23:38:53

Diagnostics Check		OK	23:32
Air Blank		0.000	23:33
Control Test		0.081	23:33
Air Blank		0.000	23:33
Subject Sample	# 1	0.000	23:34
Breath Volume		1.734	23:34
Air Blank		0.000	23:34
Air Blank		0.000	23:36
Subject Sample	# 2	0.000	23:37
Breath Volume		1.562	23:37
Air Blank		0.000	23:37
Control Test		0.081	23:38
Air Blank		0.000	23:38
Diagnostics Check		OK	23:38

Data Download Date/Time: 05/02/2019 05:12

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PYE,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: PYE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 04/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: NUNEZ-RODRIGUEZ

DATE/TIME OF TEST: 04/04/2019 00:20:09

Diagnostics Check		OK	00:14
Air Blank		0.000	00:14
Control Test		0.080	00:14
Air Blank		0.000	00:15
Subject Sample	# 1	0.045	00:15
Breath Volume		1.339	00:15
Air Blank		0.000	00:16
Air Blank		0.000	00:18
Subject Sample	# 2	0.044	00:18
Breath Volume		1.339	00:18
Air Blank		0.000	00:19
Control Test		0.077	00:19
Air Blank		0.000	00:20
Diagnostics Check		OK	00:20

Data Download Date/Time: 05/02/2019 05:03

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIEVES,CHARLIE,
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: CAMPOS,BARBARA
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 04/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: LOPEZ NATAL

DATE/TIME OF TEST: 04/06/2019 03:57:42

Diagnostics Check		OK	03:51
Air Blank		0.000	03:51
Control Test		0.078	03:52
Air Blank		0.000	03:52
Subject Sample	# 1	0.167	03:53
Breath Volume		1.203	03:53
Air Blank		0.000	03:54
Air Blank		0.000	03:55
Subject Sample	# 2	0.173	03:56
Breath Volume		1.460	03:56
Air Blank		0.000	03:56
Control Test		0.076	03:57
Air Blank		0.000	03:57
Diagnostics Check		OK	03:57

Data Download Date/Time: 05/02/2019 05:03

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAYERS,MICHELLE,H
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LOPEZ,MICHALL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 04/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: FERRAN-DORIS

DATE/TIME OF TEST: 04/06/2019 04:32:49

Diagnostics Check		OK	04:20	
Air Blank		0.000	04:21	
Control Test		0.077	04:21	
Air Blank		0.000	04:21	
Subject Sample	# 1	0.031	04:25	Volume Not Met
Breath Volume		0.906	04:25	
Air Blank		0.000	04:25	
Air Blank		0.000	04:27	
Subject Sample	# 2	0.030	04:28	
Breath Volume		1.757	04:28	
Air Blank		0.000	04:28	
Air Blank		0.000	04:30	
Subject Sample	# 3	0.030	04:31	
Breath Volume		1.523	04:31	
Air Blank		0.000	04:31	
Control Test		0.078	04:32	
Air Blank		0.000	04:32	
Diagnostics Check		OK	04:32	

Data Download Date/Time: 05/02/2019 05:03

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIEVES,CHARLIE,
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: FILALI FIKRI,NAWFAL
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 04/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: HIRACHETA

DATE/TIME OF TEST: 04/06/2019 21:38:52

Diagnostics Check		OK	21:27	
Air Blank		0.000	21:27	
Control Test		0.078	21:28	
Air Blank		0.000	21:28	
Subject Sample	# 1		21:31	No Sample Provided
Breath Volume		0.000	21:31	
Air Blank		0.000	21:32	
Air Blank		0.000	21:34	
Subject Sample	# 2	0.206	21:34	
Breath Volume		2.960	21:34	
Air Blank		0.000	21:35	
Air Blank		0.000	21:36	
Subject Sample	# 3	0.201	21:37	
Breath Volume		2.632	21:37	
Air Blank		0.000	21:37	
Control Test		0.077	21:38	
Air Blank		0.000	21:38	
Diagnostics Check		OK	21:38	

Data Download Date/Time: 05/02/2019 05:03

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCONNELL,CHARLES,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: BAEZ,BAYRON
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 04/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: ARROYO

DATE/TIME OF TEST: 04/09/2019 21:47:30

Diagnostics Check		OK	21:41
Air Blank		0.000	21:41
Control Test		0.078	21:42
Air Blank		0.000	21:42
Subject Sample	# 1	0.000	21:43
Breath Volume		1.457	21:43
Air Blank		0.000	21:43
Air Blank		0.000	21:45
Subject Sample	# 2	0.000	21:46
Breath Volume		1.527	21:46
Air Blank		0.000	21:46
Control Test		0.076	21:46
Air Blank		0.000	21:47
Diagnostics Check		OK	21:47

Data Download Date/Time: 05/02/2019 05:03

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PYE,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: PYE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 04/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: ARGUELLES MEDINA

DATE/TIME OF TEST: 04/10/2019 00:36:39

Diagnostics Check		OK	00:30
Air Blank		0.000	00:30
Control Test		0.078	00:31
Air Blank		0.000	00:31
Subject Sample	# 1	0.082	00:32
Breath Volume		1.488	00:32
Air Blank		0.000	00:32
Air Blank		0.000	00:34
Subject Sample	# 2	0.084	00:35
Breath Volume		1.550	00:35
Air Blank		0.000	00:35
Control Test		0.079	00:36
Air Blank		0.000	00:36
Diagnostics Check		OK	00:36

Data Download Date/Time: 05/02/2019 05:03

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AKINS,JUSTIN,M
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LOPEZ,MICHALL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 04/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: ESPINOSA-BRAVO

DATE/TIME OF TEST: 04/11/2019 03:28:43

Diagnostics Check		OK	03:21
Air Blank		0.000	03:22
Control Test		0.077	03:22
Air Blank		0.000	03:23
Subject Sample	# 1	0.064	03:23
Breath Volume		1.187	03:23
Air Blank		0.000	03:24
Air Blank		0.000	03:26
Subject Sample	# 2	0.065	03:27
Breath Volume		2.101	03:27
Air Blank		0.000	03:27
Control Test		0.078	03:28
Air Blank		0.000	03:28
Diagnostics Check		OK	03:28

Data Download Date/Time: 05/02/2019 05:03

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MINERVINO,KRISTIN,
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: MINERVINO,KRISTIN
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 04/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 14:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:39

SUBJECT'S LAST NAME: CATENA

DATE/TIME OF TEST: 04/12/2019 15:17:49

Diagnostics Check		OK	15:11
Air Blank		0.000	15:12
Control Test		0.079	15:12
Air Blank		0.000	15:12
Subject Sample	# 1	0.165	15:13
Breath Volume		3.593	15:13
Air Blank		0.000	15:14
Air Blank		0.000	15:15
Subject Sample	# 2	0.158	15:16
Breath Volume		2.464	15:16
Air Blank		0.000	15:16
Control Test		0.076	15:17
Air Blank		0.000	15:17
Diagnostics Check		OK	15:17

Data Download Date/Time: 05/02/2019 05:03

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAYERS,MICHELLE,H
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: SAYERS,MICHELLE
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 04/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: SINGH

DATE/TIME OF TEST: 04/20/2019 01:25:00

Diagnostics Check		OK	01:17
Air Blank		0.000	01:18
Control Test		0.079	01:18
Air Blank		0.000	01:19
Subject Sample	# 1	0.236	01:20
Breath Volume		2.433	01:20
Air Blank		0.000	01:21
Air Blank		0.000	01:22
Subject Sample	# 2	0.226	01:23
Breath Volume		2.207	01:23
Air Blank		0.000	01:24
Control Test		0.077	01:24
Air Blank		0.000	01:24
Diagnostics Check		OK	01:25

Data Download Date/Time: 05/02/2019 05:03

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,DANIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: AKINS,JUSTIN
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 04/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: DILL

DATE/TIME OF TEST: 04/21/2019 00:27:49

Diagnostics Check		OK	00:21
Air Blank		0.000	00:21
Control Test		0.078	00:22
Air Blank		0.000	00:22
Subject Sample	# 1	0.239	00:23
Breath Volume		2.472	00:23
Air Blank		0.000	00:23
Air Blank		0.000	00:25
Subject Sample	# 2	0.247	00:26
Breath Volume		1.796	00:26
Air Blank		0.000	00:26
Control Test		0.076	00:27
Air Blank		0.000	00:27
Diagnostics Check		OK	00:27

Data Download Date/Time: 05/02/2019 05:03

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSELL,ROBERT,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: BROOKS,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: LUCAS

DATE/TIME OF TEST: 04/23/2019 00:30:54

Diagnostics Check		OK	00:24	
Air Blank		0.000	00:25	
Control Test		0.076	00:25	
Air Blank		0.000	00:26	
Subject Sample	# 1	0.149	00:26	
Breath Volume		1.441	00:26	
Air Blank		0.000	00:27	
Air Blank		0.000	00:28	
Subject Sample	# 2		00:29	Slope Not Met
Breath Volume		0.664	00:29	
Air Blank		0.000	00:29	
Control Test		0.075	00:30	
Air Blank		0.000	00:30	
Diagnostics Check		OK	00:30	

Data Download Date/Time: 05/02/2019 05:03

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PYE,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: PYE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 04/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:35

SUBJECT'S LAST NAME: DIAZ FIGUEROA

DATE/TIME OF TEST: 04/24/2019 06:05:09

Diagnostics Check		OK	05:58
Air Blank		0.000	05:59
Control Test		0.079	05:59
Air Blank		0.000	05:59
Subject Sample	# 1	0.000	06:00
Breath Volume		2.062	06:00
Air Blank		0.000	06:00
Air Blank		0.000	06:03
Subject Sample	# 2	0.000	06:03
Breath Volume		1.250	06:03
Air Blank		0.000	06:04
Control Test		0.076	06:04
Air Blank		0.000	06:05
Diagnostics Check		OK	06:05

Data Download Date/Time: 05/02/2019 05:03

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAYERS,MICHELLE,H
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: SAYERS,MICHELLE
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 04/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: KIMRAJ

DATE/TIME OF TEST: 04/25/2019 02:13:46

Diagnostics Check		OK	02:07
Air Blank		0.000	02:07
Control Test		0.079	02:08
Air Blank		0.000	02:08
Subject Sample	# 1	0.097	02:09
Breath Volume		1.687	02:09
Air Blank		0.000	02:09
Air Blank		0.000	02:11
Subject Sample	# 2	0.091	02:12
Breath Volume		1.585	02:12
Air Blank		0.000	02:12
Control Test		0.079	02:13
Air Blank		0.000	02:13
Diagnostics Check		OK	02:13

Data Download Date/Time: 05/02/2019 05:03

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AKINS,JUSTIN,M
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LOPEZ,MICHALL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 04/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: AVILEZ CANO

DATE/TIME OF TEST: 04/25/2019 03:11:37

Diagnostics Check		OK	03:04
Air Blank		0.000	03:05
Control Test		0.079	03:05
Air Blank		0.000	03:06
Subject Sample	# 1	0.195	03:06
Breath Volume		1.660	03:06
Air Blank		0.000	03:07
Air Blank		0.000	03:09
Subject Sample	# 2	0.189	03:10
Breath Volume		1.441	03:10
Air Blank		0.000	03:10
Control Test		0.079	03:11
Air Blank		0.000	03:11
Diagnostics Check		OK	03:11

Data Download Date/Time: 05/02/2019 05:03

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORD,GLENN,W
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: FORD,GLENN
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 04/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 15:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:30

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 04/27/2019 15:58:58

Diagnostics Check		OK	15:53	
Air Blank		0.000	15:53	
Control Test		0.080	15:54	
Air Blank		0.000	15:54	
Subject Sample	# 1		15:57	No Sample Provided
Breath Volume		0.000	15:57	
Air Blank		0.000	15:58	
Air Blank		0.000	15:58	

Data Download Date/Time: 05/02/2019 05:03

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PYE,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: PYE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 04/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 04/27/2019 22:26:41

Diagnostics Check		OK	22:20
Air Blank		0.000	22:21
Control Test		0.079	22:21
Air Blank		0.000	22:21
Subject Sample	# 1	0.176	22:22
Breath Volume		2.382	22:22
Air Blank		0.000	22:22
Air Blank		0.000	22:24
Subject Sample	# 2	0.177	22:25
Breath Volume		2.433	22:25
Air Blank		0.000	22:25
Control Test		0.079	22:26
Air Blank		0.000	22:26
Diagnostics Check		OK	22:26

Data Download Date/Time: 05/02/2019 05:03

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,DANIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: MINERVINO/BURROUGHS,KRISTIN/KORTNEY
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 04/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 06:02
GENDER: Male

OBSERVATION PERIOD BEGAN: 06:34

SUBJECT'S LAST NAME: CAICEDO

DATE/TIME OF TEST: 04/28/2019 07:09:40

Diagnostics Check		OK	06:55	
Air Blank		0.000	06:55	
Control Test		0.079	06:55	
Air Blank		0.000	06:56	
Subject Sample	# 1	0.197	06:58	
Breath Volume		1.566	06:58	
Air Blank		0.000	06:59	
Air Blank		0.000	07:00	
Subject Sample	# 2	0.179	07:04	Volume Not Met
Breath Volume		0.906	07:04	
Air Blank		0.000	07:04	
Air Blank		0.000	07:06	
Subject Sample	# 3	0.192	07:08	
Breath Volume		1.898	07:08	
Air Blank		0.000	07:08	
Control Test		0.076	07:09	
Air Blank		0.000	07:09	
Diagnostics Check		OK	07:09	

Data Download Date/Time: 05/02/2019 05:03

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,LANDON,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WALKER,LANDON
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 04/02/2019 00:24:21

Diagnostics Check		OK	00:16
Air Blank		0.000	00:17
Control Test		0.082	00:17
Air Blank		0.000	00:18
Subject Sample	# 1	0.171	00:19
Breath Volume		1.464	00:19
Air Blank		0.000	00:20
Air Blank		0.000	00:22
Subject Sample	# 2	0.178	00:22
Breath Volume		1.207	00:22
Air Blank		0.000	00:23
Control Test		0.078	00:23
Air Blank		0.000	00:24
Diagnostics Check		OK	00:24

Data Download Date/Time: 04/04/2019 09:37

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRISTOL,JARRET,W
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: RAGUSA,JASON
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: HUNT

DATE/TIME OF TEST: 04/02/2019 23:41:10

Diagnostics Check		OK	23:35
Air Blank		0.000	23:35
Control Test		0.082	23:35
Air Blank		0.000	23:36
Subject Sample	# 1	0.000	23:36
Breath Volume		2.835	23:36
Air Blank		0.000	23:37
Air Blank		0.000	23:39
Subject Sample	# 2	0.000	23:39
Breath Volume		2.757	23:39
Air Blank		0.000	23:40
Control Test		0.080	23:40
Air Blank		0.000	23:41
Diagnostics Check		OK	23:41

Data Download Date/Time: 04/04/2019 09:37

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RACER,JACOB,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: RACER,JACOB
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: MALLETTE

DATE/TIME OF TEST: 04/03/2019 02:49:59

Diagnostics Check		OK	02:41	
Air Blank		0.000	02:41	
Control Test		0.082	02:42	
Air Blank		0.000	02:42	
Subject Sample	# 1	0.239	02:45	Volume Not Met
Breath Volume		0.312	02:45	
Air Blank		0.000	02:46	
Air Blank		0.000	02:48	
Subject Sample	# 2		02:48	Subject Test Refused
Breath Volume		0.000	02:48	
Air Blank		0.000	02:49	
Control Test		0.076	02:49	
Air Blank		0.000	02:49	
Diagnostics Check		OK	02:49	

Data Download Date/Time: 04/04/2019 09:37

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RACER,JACOB,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: RACER,JACOB
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: MALLETTE

DATE/TIME OF TEST: 04/03/2019 03:07:03

Diagnostics Check		OK	02:57
Air Blank		0.000	02:58
Control Test		0.081	02:58
Air Blank		0.000	02:59
Subject Sample	# 1	0.233	03:00
Breath Volume		2.324	03:00
Air Blank		0.000	03:01
Air Blank		0.000	03:02
Subject Sample	# 2	0.224	03:05
Breath Volume		1.746	03:05
Air Blank		0.000	03:06
Control Test		0.078	03:06
Air Blank		0.000	03:06
Diagnostics Check		OK	03:07

Data Download Date/Time: 04/04/2019 09:37

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDE,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WILDE,MATTHEW
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 16:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:04

SUBJECT'S LAST NAME: BYRNE

DATE/TIME OF TEST: 04/04/2019 17:28:31

Diagnostics Check		OK	17:25	
Air Blank		0.000	17:26	
Control Test		0.080	17:26	
Air Blank		0.000	17:26	
Subject Sample	# 1		17:27	Subject Test Refused
Breath Volume		0.000	17:27	
Air Blank		0.000	17:27	
Control Test		0.081	17:27	
Air Blank		0.000	17:28	
Diagnostics Check		OK	17:28	

Data Download Date/Time: 05/09/2019 09:36

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOGGAN,MARLON,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BOGGAN,MARLON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 04/05/2019 22:31:11

Diagnostics Check		OK	22:25
Air Blank		0.000	22:25
Control Test		0.081	22:26
Air Blank		0.000	22:26
Subject Sample	# 1	0.233	22:26
Breath Volume		1.539	22:26
Air Blank		0.000	22:27
Air Blank		0.000	22:29
Subject Sample	# 2	0.221	22:29
Breath Volume		1.464	22:29
Air Blank		0.000	22:30
Control Test		0.080	22:30
Air Blank		0.000	22:31
Diagnostics Check		OK	22:31

Data Download Date/Time: 05/09/2019 09:36

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,LANDON,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WALKER,LANDON
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: TATE

DATE/TIME OF TEST: 04/07/2019 03:24:11

Diagnostics Check		OK	03:20	
Air Blank		0.000	03:21	
Control Test		0.080	03:21	
Air Blank		0.000	03:22	
Subject Sample	# 1		03:22	Subject Test Refused
Breath Volume		0.000	03:22	
Air Blank		0.000	03:23	
Control Test		0.081	03:23	
Air Blank		0.000	03:24	
Diagnostics Check		OK	03:24	

Data Download Date/Time: 05/09/2019 09:36

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FITCHETT,ADAM,R
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SWAIN,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: LISTER

DATE/TIME OF TEST: 04/07/2019 04:30:51

Diagnostics Check		OK	04:24
Air Blank		0.000	04:24
Control Test		0.081	04:25
Air Blank		0.000	04:25
Subject Sample	# 1	0.250	04:26
Breath Volume		2.261	04:26
Air Blank		0.000	04:26
Air Blank		0.000	04:28
Subject Sample	# 2	0.261	04:29
Breath Volume		2.554	04:29
Air Blank		0.000	04:29
Control Test		0.080	04:30
Air Blank		0.000	04:30
Diagnostics Check		OK	04:30

Data Download Date/Time: 05/09/2019 09:36

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRANE,RICHARD,B
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: DRANE,RICHARD
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 04:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:42

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 04/07/2019 06:16:35

Diagnostics Check		OK	06:04	
Air Blank		0.000	06:05	
Control Test		0.081	06:05	
Air Blank		0.000	06:06	
Subject Sample	# 1	0.189	06:09	Volume Not Met
Breath Volume		0.035	06:09	
Air Blank		0.000	06:10	
Air Blank		0.000	06:11	
Subject Sample	# 2	0.182	06:14	Volume Not Met
Breath Volume		0.558	06:14	
Air Blank		0.000	06:15	
Control Test		0.078	06:15	
Air Blank		0.000	06:16	
Diagnostics Check		OK	06:16	

Data Download Date/Time: 05/09/2019 09:36

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRANE,RICHARD,B
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: DRANE,RICHARD
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 04:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:42

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 04/07/2019 06:25:58

Diagnostics Check		OK	06:19
Air Blank		0.000	06:19
Control Test		0.079	06:19
Air Blank		0.000	06:20
Subject Sample	# 1	0.204	06:21
Breath Volume		1.914	06:21
Air Blank		0.000	06:21
Air Blank		0.000	06:23
Subject Sample	# 2	0.185	06:24
Breath Volume		1.144	06:24
Air Blank		0.000	06:25
Control Test		0.079	06:25
Air Blank		0.000	06:25
Diagnostics Check		OK	06:25

Data Download Date/Time: 05/09/2019 09:36

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROOKS,ROYCE,G
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LANGER,MATTHEW
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 10:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:27

SUBJECT'S LAST NAME: BRAZZELL

DATE/TIME OF TEST: 04/08/2019 11:01:44

Diagnostics Check		OK	10:54
Air Blank		0.000	10:54
Control Test		0.081	10:55
Air Blank		0.000	10:55
Subject Sample	# 1	0.000	10:57
Breath Volume		1.867	10:57
Air Blank		0.000	10:57
Air Blank		0.000	10:59
Subject Sample	# 2	0.000	11:00
Breath Volume		2.062	11:00
Air Blank		0.000	11:00
Control Test		0.081	11:01
Air Blank		0.000	11:01
Diagnostics Check		OK	11:01

Data Download Date/Time: 05/09/2019 09:36

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,LANDON,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: RAGUSA,NICHOLE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: SCHAEFER

DATE/TIME OF TEST: 04/10/2019 22:51:00

Diagnostics Check		OK	22:47	
Air Blank		0.000	22:48	
Control Test		0.081	22:48	
Air Blank		0.000	22:49	
Subject Sample	# 1		22:49	Subject Test Refused
Breath Volume		0.000	22:49	
Air Blank		0.000	22:50	
Control Test		0.080	22:50	
Air Blank		0.000	22:50	
Diagnostics Check		OK	22:51	

Data Download Date/Time: 05/09/2019 09:36

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,RAY,C
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: HAMILTON,RAY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 10:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:44

SUBJECT'S LAST NAME: HOLDNAK

DATE/TIME OF TEST: 04/11/2019 12:12:36

Diagnostics Check		OK	12:06
Air Blank		0.000	12:06
Control Test		0.080	12:07
Air Blank		0.000	12:07
Subject Sample	# 1	0.000	12:07
Breath Volume		1.531	12:07
Air Blank		0.000	12:08
Air Blank		0.000	12:10
Subject Sample	# 2	0.000	12:11
Breath Volume		3.238	12:11
Air Blank		0.000	12:11
Control Test		0.080	12:12
Air Blank		0.000	12:12
Diagnostics Check		OK	12:12

Data Download Date/Time: 05/09/2019 09:36

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROOKS,ROYCE,G
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: ARMITAGE,MATTHEW
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 07:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:21

SUBJECT'S LAST NAME: COLSON

DATE/TIME OF TEST: 04/12/2019 07:59:52

Diagnostics Check		OK	07:51
Air Blank		0.000	07:51
Control Test		0.081	07:52
Air Blank		0.000	07:52
Subject Sample	# 1	0.193	07:53
Breath Volume		1.179	07:53
Air Blank		0.000	07:54
Air Blank		0.000	07:56
Subject Sample	# 2	0.182	07:58
Breath Volume		1.347	07:58
Air Blank		0.000	07:58
Control Test		0.080	07:59
Air Blank		0.000	07:59
Diagnostics Check		OK	07:59

Data Download Date/Time: 05/09/2019 09:36

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIMBRIL,JOSHUA,B
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: KIMBRIL,JOSHUA
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: TRIESCHMANN

DATE/TIME OF TEST: 04/13/2019 00:55:42

Diagnostics Check		OK	00:48
Air Blank		0.000	00:49
Control Test		0.080	00:49
Air Blank		0.000	00:50
Subject Sample	# 1	0.065	00:51
Breath Volume		1.675	00:51
Air Blank		0.000	00:51
Air Blank		0.000	00:53
Subject Sample	# 2	0.066	00:54
Breath Volume		1.402	00:54
Air Blank		0.000	00:54
Control Test		0.081	00:55
Air Blank		0.000	00:55
Diagnostics Check		OK	00:55

Data Download Date/Time: 05/09/2019 09:36

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOGGAN,MARLON,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BOGGAN,MARLON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: HANZLICK

DATE/TIME OF TEST: 04/13/2019 01:36:31

Diagnostics Check		OK	01:25	
Air Blank		0.000	01:25	
Control Test		0.080	01:25	
Air Blank		0.000	01:26	
Subject Sample	# 1		01:29	No Sample Provided
Breath Volume		0.000	01:29	
Air Blank		0.000	01:30	
Air Blank		0.000	01:32	
Subject Sample	# 2		01:35	No Sample Provided
Breath Volume		0.000	01:35	
Air Blank		0.000	01:35	
Control Test		0.081	01:35	
Air Blank		0.000	01:36	
Diagnostics Check		OK	01:36	

Data Download Date/Time: 05/09/2019 09:36

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREIRA,JOSE,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: PEREIRA,JOSE
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: HUFF

DATE/TIME OF TEST: 04/13/2019 01:49:51

Diagnostics Check		OK	01:42
Air Blank		0.000	01:42
Control Test		0.081	01:43
Air Blank		0.000	01:43
Subject Sample	# 1	0.132	01:44
Breath Volume		1.554	01:44
Air Blank		0.000	01:45
Air Blank		0.000	01:47
Subject Sample	# 2	0.124	01:48
Breath Volume		2.050	01:48
Air Blank		0.000	01:48
Control Test		0.079	01:49
Air Blank		0.000	01:49
Diagnostics Check		OK	01:49

Data Download Date/Time: 05/09/2019 09:37

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYBRAND,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LYBRAND,CHRISTOPHER
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 04:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:35

SUBJECT'S LAST NAME: WADDELL

DATE/TIME OF TEST: 04/16/2019 06:04:34

Diagnostics Check		OK	05:57
Air Blank		0.000	05:57
Control Test		0.081	05:58
Air Blank		0.000	05:58
Subject Sample	# 1	0.167	05:59
Breath Volume		1.417	05:59
Air Blank		0.000	05:59
Air Blank		0.000	06:01
Subject Sample	# 2	0.168	06:02
Breath Volume		1.925	06:02
Air Blank		0.000	06:03
Control Test		0.081	06:03
Air Blank		0.000	06:04
Diagnostics Check		OK	06:04

Data Download Date/Time: 05/09/2019 09:37

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICE,DUSTAN,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: RICE,DUSTAN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 20:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:02

SUBJECT'S LAST NAME: ROSAMOND

DATE/TIME OF TEST: 04/18/2019 21:29:20

Diagnostics Check		OK	21:26	
Air Blank		0.000	21:27	
Control Test		0.081	21:27	
Air Blank		0.000	21:27	
Subject Sample	# 1		21:27	Subject Test Refused
Breath Volume		0.000	21:27	
Air Blank		0.000	21:28	
Control Test		0.081	21:28	
Air Blank		0.000	21:29	
Diagnostics Check		OK	21:29	

Data Download Date/Time: 05/09/2019 09:37

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIS,ROBERT,B
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WILLIS,ROBERT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: DUPERON

DATE/TIME OF TEST: 04/20/2019 03:21:13

Diagnostics Check		OK	03:14
Air Blank		0.000	03:15
Control Test		0.081	03:15
Air Blank		0.000	03:16
Subject Sample	# 1	0.153	03:16
Breath Volume		1.218	03:16
Air Blank		0.000	03:17
Air Blank		0.000	03:19
Subject Sample	# 2	0.155	03:19
Breath Volume		2.125	03:19
Air Blank		0.000	03:20
Control Test		0.080	03:20
Air Blank		0.000	03:21
Diagnostics Check		OK	03:21

Data Download Date/Time: 05/09/2019 09:37

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERSON,JOHN,C
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SANDERSON,JOHN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: SOSA

DATE/TIME OF TEST: 04/20/2019 04:23:23

Diagnostics Check		OK	04:15
Air Blank		0.000	04:16
Control Test		0.082	04:16
Air Blank		0.000	04:17
Subject Sample	# 1	0.219	04:18
Breath Volume		1.859	04:18
Air Blank		0.000	04:19
Air Blank		0.000	04:21
Subject Sample	# 2	0.219	04:21
Breath Volume		2.207	04:21
Air Blank		0.000	04:22
Control Test		0.078	04:22
Air Blank		0.000	04:23
Diagnostics Check		OK	04:23

Data Download Date/Time: 05/09/2019 09:37

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,RAY,C
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: HAMILTON,RAY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 06:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:00

SUBJECT'S LAST NAME: BONNER

DATE/TIME OF TEST: 04/21/2019 07:25:10

Diagnostics Check		OK	07:21	
Air Blank		0.000	07:21	
Control Test		0.081	07:22	
Air Blank		0.000	07:22	
Subject Sample	# 1		07:24	Sequence Aborted
Breath Volume		0.000	07:24	
Air Blank		0.000	07:25	

Data Download Date/Time: 05/09/2019 09:37

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,RAY,C
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: HAMILTON,RAY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 06:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:03

SUBJECT'S LAST NAME: BONNER

DATE/TIME OF TEST: 04/21/2019 07:27:55

Diagnostics Check		OK	07:25	
Air Blank		0.000	07:25	
Control Test		0.082	07:25	
Air Blank		0.000	07:26	
Subject Sample	# 1		07:26	Subject Test Refused
Breath Volume		0.000	07:26	
Air Blank		0.000	07:27	
Control Test		0.082	07:27	
Air Blank		0.000	07:27	
Diagnostics Check		OK	07:27	

Data Download Date/Time: 05/09/2019 09:37

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MLYNARSKI,NIKLAS,E
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: MLYNARSKI,NIKLAS
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: ZURICK

DATE/TIME OF TEST: 04/24/2019 01:25:22

Diagnostics Check		OK	01:22	
Air Blank		0.000	01:22	
Control Test		0.079	01:23	
Air Blank		0.000	01:23	
Subject Sample	# 1		01:24	Subject Test Refused
Breath Volume		0.000	01:24	
Air Blank		0.000	01:24	
Control Test		0.080	01:24	
Air Blank		0.000	01:25	
Diagnostics Check		OK	01:25	

Data Download Date/Time: 05/09/2019 09:37

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RACER,JACOB,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: RACER,JACOB
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: MCCOMB

DATE/TIME OF TEST: 04/25/2019 03:51:43

Diagnostics Check		OK	03:40	
Air Blank		0.000	03:40	
Control Test		0.081	03:41	
Air Blank		0.000	03:41	
Subject Sample	# 1		03:44	No Sample Provided
Breath Volume		0.000	03:44	
Air Blank		0.000	03:45	
Air Blank		0.000	03:47	
Subject Sample	# 2		03:50	No Sample Provided
Breath Volume		0.000	03:50	
Air Blank		0.000	03:50	
Control Test		0.080	03:51	
Air Blank		0.000	03:51	
Diagnostics Check		OK	03:51	

Data Download Date/Time: 05/09/2019 09:37

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIS,ROBERT,B
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: DUS,CHRISTINA
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: BURNETT

DATE/TIME OF TEST: 04/26/2019 05:05:45

Diagnostics Check		OK	04:58
Air Blank		0.000	04:59
Control Test		0.081	04:59
Air Blank		0.000	04:59
Subject Sample	# 1	0.086	05:00
Breath Volume		1.425	05:00
Air Blank		0.000	05:01
Air Blank		0.000	05:03
Subject Sample	# 2	0.085	05:04
Breath Volume		2.554	05:04
Air Blank		0.000	05:04
Control Test		0.079	05:05
Air Blank		0.000	05:05
Diagnostics Check		OK	05:05

Data Download Date/Time: 05/09/2019 09:37

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,RAYMOND,S
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: MILLER,RAYMOND
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 09:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:16

SUBJECT'S LAST NAME: HEMPY

DATE/TIME OF TEST: 04/26/2019 10:45:33

Diagnostics Check		OK	10:38	
Air Blank		0.000	10:38	
Control Test		0.081	10:38	
Air Blank		0.000	10:39	
Subject Sample	# 1	0.093	10:42	
Breath Volume		1.140	10:42	
Air Blank		0.000	10:43	
Air Blank			10:45	Ambient Fail
Air Blank		0.000	10:45	

Data Download Date/Time: 05/09/2019 09:37

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,RAYMOND,S
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: MILLER,RAYMOND
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 09:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:16

SUBJECT'S LAST NAME: HEMPY

DATE/TIME OF TEST: 04/26/2019 10:59:28

Diagnostics Check		OK	10:52
Air Blank		0.000	10:52
Control Test		0.081	10:52
Air Blank		0.000	10:53
Subject Sample	# 1	0.090	10:54
Breath Volume		2.289	10:54
Air Blank		0.000	10:54
Air Blank		0.000	10:56
Subject Sample	# 2	0.081	10:57
Breath Volume		2.980	10:57
Air Blank		0.000	10:58
Control Test		0.079	10:58
Air Blank		0.000	10:59
Diagnostics Check		OK	10:59

Data Download Date/Time: 05/09/2019 09:37

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDE,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WILDE,MATTHEW
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 11:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:49

SUBJECT'S LAST NAME: CASTANEDASANDOVAL

DATE/TIME OF TEST: 04/27/2019 12:15:49

Diagnostics Check		OK	12:09
Air Blank		0.000	12:09
Control Test		0.081	12:10
Air Blank		0.000	12:10
Subject Sample	# 1	0.273	12:11
Breath Volume		1.750	12:11
Air Blank		0.000	12:11
Air Blank		0.000	12:13
Subject Sample	# 2	0.290	12:14
Breath Volume		2.722	12:14
Air Blank		0.000	12:14
Control Test		0.080	12:15
Air Blank		0.000	12:15
Diagnostics Check		OK	12:15

Data Download Date/Time: 05/09/2019 09:37

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIMBRIL,JOSHUA,B
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SPENCE,JUSTIN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 21:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:37

SUBJECT'S LAST NAME: SANDERS

DATE/TIME OF TEST: 04/27/2019 23:04:30

Diagnostics Check		OK	22:57
Air Blank		0.000	22:57
Control Test		0.081	22:58
Air Blank		0.000	22:58
Subject Sample	# 1	0.081	22:59
Breath Volume		1.582	22:59
Air Blank		0.000	22:59
Air Blank		0.000	23:01
Subject Sample	# 2	0.083	23:02
Breath Volume		2.398	23:02
Air Blank		0.000	23:03
Control Test		0.081	23:03
Air Blank		0.000	23:04
Diagnostics Check		OK	23:04

Data Download Date/Time: 05/09/2019 09:37

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROOKS,ROYCE,G
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: ABEL,ANTHONY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 09:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:35

SUBJECT'S LAST NAME: HENDERSON

DATE/TIME OF TEST: 04/28/2019 11:01:36

Diagnostics Check		OK	10:55
Air Blank		0.000	10:55
Control Test		0.081	10:55
Air Blank		0.000	10:56
Subject Sample	# 1	0.000	10:57
Breath Volume		2.054	10:57
Air Blank		0.000	10:57
Air Blank		0.000	10:59
Subject Sample	# 2	0.000	11:00
Breath Volume		1.347	11:00
Air Blank		0.000	11:00
Control Test		0.081	11:01
Air Blank		0.000	11:01
Diagnostics Check		OK	11:01

Data Download Date/Time: 05/09/2019 09:37

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILHELM,TONY,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WILHELM,TONY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 11:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:00

SUBJECT'S LAST NAME: SIMIEN

DATE/TIME OF TEST: 04/28/2019 12:26:47

Diagnostics Check		OK	12:23	
Air Blank		0.000	12:24	
Control Test		0.081	12:24	
Air Blank		0.000	12:24	
Subject Sample	# 1		12:25	Subject Test Refused
Breath Volume		0.000	12:25	
Air Blank		0.000	12:25	
Control Test		0.081	12:26	
Air Blank		0.000	12:26	
Diagnostics Check		OK	12:26	

Data Download Date/Time: 05/09/2019 09:37

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,LANCE,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: CHASE,DUSTIN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 08:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:20

SUBJECT'S LAST NAME: KOONCE

DATE/TIME OF TEST: 04/29/2019 09:46:42

Diagnostics Check		OK	09:40
Air Blank		0.000	09:41
Control Test		0.081	09:41
Air Blank		0.000	09:42
Subject Sample	# 1	0.000	09:42
Breath Volume		2.542	09:42
Air Blank		0.000	09:42
Air Blank		0.000	09:44
Subject Sample	# 2	0.000	09:45
Breath Volume		3.003	09:45
Air Blank		0.000	09:45
Control Test		0.080	09:46
Air Blank		0.000	09:46
Diagnostics Check		OK	09:46

Data Download Date/Time: 05/09/2019 09:37

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MONCION,C
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: PEREZ HERNANDEZ

DATE/TIME OF TEST: 04/01/2019 00:12:15

Diagnostics Check		OK	00:05
Air Blank		0.000	00:06
Control Test		0.083	00:06
Air Blank		0.000	00:07
Subject Sample	# 1	0.212	00:07
Breath Volume		2.160	00:07
Air Blank		0.000	00:08
Air Blank		0.000	00:10
Subject Sample	# 2	0.218	00:10
Breath Volume		2.117	00:10
Air Blank		0.000	00:11
Control Test		0.083	00:11
Air Blank		0.000	00:12
Diagnostics Check		OK	00:12

Data Download Date/Time: 04/04/2019 17:04

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005248

INSTRUMENT REGISTERED TO: FISH & WILDLIFE CC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,WILLIAM,J
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: THOMPSON,WILLIAM
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 620051
DRY GAS EXPIRY DATE: 11/05/2019
VIOLATION CODE: BUI
ARREST TIME: 15:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:39

SUBJECT'S LAST NAME: CORARINSKY

DATE/TIME OF TEST: 04/07/2019 16:02:39

Diagnostics Check		OK	15:59	
Air Blank		0.000	16:00	
Control Test		0.084	16:00	
Air Blank		0.000	16:01	
Subject Sample	# 1		16:01	Subject Test Refused
Breath Volume		0.000	16:01	
Air Blank		0.000	16:01	
Control Test		0.083	16:02	
Air Blank		0.000	16:02	
Diagnostics Check		OK	16:02	

Data Download Date/Time: 04/30/2019 23:42

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005248

INSTRUMENT REGISTERED TO: FISH & WILDLIFE CC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRAWINSKI,MARCIN,A
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: SAMPLE,CARSON
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 620051
DRY GAS EXPIRY DATE: 11/05/2019
VIOLATION CODE: BUI
ARREST TIME: 16:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:12

SUBJECT'S LAST NAME: ARAGON

DATE/TIME OF TEST: 04/07/2019 16:48:47

Diagnostics Check		OK	16:33	
Air Blank		0.000	16:33	
Control Test		0.083	16:34	
Air Blank		0.000	16:34	
Subject Sample	# 1	0.080	16:37	
Breath Volume		2.121	16:37	
Air Blank		0.000	16:38	
Air Blank		0.000	16:40	
Subject Sample	# 2	0.069	16:43	Volume Not Met
Breath Volume		0.125	16:43	
Air Blank		0.000	16:44	
Air Blank		0.000	16:45	
Subject Sample	# 3	0.079	16:46	
Breath Volume		2.332	16:46	
Air Blank		0.000	16:47	
Control Test		0.080	16:48	
Air Blank		0.000	16:48	
Diagnostics Check		OK	16:48	

Data Download Date/Time: 04/30/2019 23:42

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005248

INSTRUMENT REGISTERED TO: FISH & WILDLIFE CC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,ADRIAN,
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: PEREZ,ADRIAN
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 620051
DRY GAS EXPIRY DATE: 11/05/2019
VIOLATION CODE: BUI
ARREST TIME: 16:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:58

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 04/07/2019 17:42:37

Diagnostics Check		OK	17:39	
Air Blank		0.000	17:39	
Control Test		0.082	17:39	
Air Blank		0.000	17:40	
Subject Sample	# 1		17:41	Subject Test Refused
Breath Volume		0.000	17:41	
Air Blank		0.000	17:41	
Control Test		0.083	17:42	
Air Blank		0.000	17:42	
Diagnostics Check		OK	17:42	

Data Download Date/Time: 04/30/2019 23:42

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005248

INSTRUMENT REGISTERED TO: FISH & WILDLIFE CC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOPPE,ZACHARY,D
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: HEDGEPETH,AUSTIN
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 620051
DRY GAS EXPIRY DATE: 11/05/2019
VIOLATION CODE: BUI
ARREST TIME: 18:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:34

SUBJECT'S LAST NAME: RADA

DATE/TIME OF TEST: 04/07/2019 19:02:40

Diagnostics Check		OK	18:58	
Air Blank		0.000	18:58	
Control Test		0.081	18:59	
Air Blank		0.000	18:59	
Subject Sample	# 1		19:01	Subject Test Refused
Breath Volume		0.000	19:01	
Air Blank		0.000	19:01	
Control Test		0.083	19:02	
Air Blank		0.000	19:02	
Diagnostics Check		OK	19:02	

Data Download Date/Time: 04/30/2019 23:42

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005248

INSTRUMENT REGISTERED TO: FISH & WILDLIFE CC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REYNAUD,GEORGE,
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: OSORIO,MILTON
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 620051
DRY GAS EXPIRY DATE: 11/05/2019
VIOLATION CODE: BUI
ARREST TIME: 18:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: HERNANDEZ CRUZ

DATE/TIME OF TEST: 04/07/2019 19:21:12

Diagnostics Check		OK	19:14
Air Blank		0.000	19:15
Control Test		0.083	19:15
Air Blank		0.000	19:16
Subject Sample	# 1	0.189	19:16
Breath Volume		2.402	19:16
Air Blank		0.000	19:17
Air Blank		0.000	19:19
Subject Sample	# 2	0.181	19:19
Breath Volume		2.253	19:19
Air Blank		0.000	19:20
Control Test		0.083	19:20
Air Blank		0.000	19:21
Diagnostics Check		OK	19:21

Data Download Date/Time: 04/30/2019 23:42

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005248

INSTRUMENT REGISTERED TO: FISH & WILDLIFE CC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REYNAUD,GEORGE,
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: MOSCHIANO,DAVID
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 620051
DRY GAS EXPIRY DATE: 11/05/2019
VIOLATION CODE: BUI
ARREST TIME: 18:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:01

SUBJECT'S LAST NAME: YSALGUEZ

DATE/TIME OF TEST: 04/07/2019 19:30:13

Diagnostics Check		OK	19:26	
Air Blank		0.000	19:26	
Control Test		0.082	19:26	
Air Blank		0.000	19:27	
Subject Sample	# 1		19:28	Subject Test Refused
Breath Volume		0.000	19:28	
Air Blank		0.000	19:29	
Control Test		0.082	19:29	
Air Blank		0.000	19:30	
Diagnostics Check		OK	19:30	

Data Download Date/Time: 04/30/2019 23:42

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005248

INSTRUMENT REGISTERED TO: FISH & WILDLIFE CC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DESPIAN,MIGUEL,M
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: DESPIAN,MIGUEL
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 620051
DRY GAS EXPIRY DATE: 11/05/2019
VIOLATION CODE: BUI
ARREST TIME: 20:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 04/28/2019 21:36:12

Diagnostics Check	OK	21:34	
Air Blank	0.000	21:34	
Control Test	0.082	21:35	
Air Blank	0.000	21:35	
		21:35	Sequence Aborted
Air Blank	0.000	21:36	

Data Download Date/Time: 04/30/2019 23:42

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005248

INSTRUMENT REGISTERED TO: FISH & WILDLIFE CC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DESPIAN,MIGUEL,M
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: DESPIAN,MIGUEL
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 620051
DRY GAS EXPIRY DATE: 11/05/2019
VIOLATION CODE: BUI
ARREST TIME: 20:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 04/28/2019 21:42:22

Diagnostics Check		OK	21:39	
Air Blank		0.000	21:39	
Control Test		0.082	21:40	
Air Blank		0.000	21:40	
Subject Sample	# 1		21:40	Subject Test Refused
Breath Volume		0.000	21:40	
Air Blank		0.000	21:41	
Control Test		0.082	21:41	
Air Blank		0.000	21:42	
Diagnostics Check		OK	21:42	

Data Download Date/Time: 04/30/2019 23:42

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CEPHUS,JAMIN,I
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: SCHRAM,AMANDA
ARRESTING OFFICER AGENCY: Taylor County SO
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 13:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:19

SUBJECT'S LAST NAME: DEMOCH

DATE/TIME OF TEST: 04/01/2019 16:46:42

Diagnostics Check		OK	16:39
Air Blank		0.000	16:40
Control Test		0.080	16:40
Air Blank		0.000	16:40
Subject Sample	# 1	0.234	16:41
Breath Volume		1.503	16:41
Air Blank		0.000	16:42
Air Blank		0.000	16:44
Subject Sample	# 2	0.239	16:45
Breath Volume		1.710	16:45
Air Blank		0.000	16:45
Control Test		0.079	16:46
Air Blank		0.000	16:46
Diagnostics Check		OK	16:46

Data Download Date/Time: 04/30/2019 13:46

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEYER,CHARLES,L
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: BROWN,STEVEN
ARRESTING OFFICER AGENCY: Perry PD
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: HAMILTON

DATE/TIME OF TEST: 04/18/2019 00:58:45

Diagnostics Check		OK	00:50
Air Blank		0.000	00:50
Control Test		0.080	00:50
Air Blank		0.000	00:51
Subject Sample	# 1	0.320	00:52
Breath Volume		2.093	00:52
Air Blank		0.000	00:53
Air Blank		0.000	00:55
Subject Sample	# 2	0.315	00:57
Breath Volume		1.589	00:57
Air Blank		0.000	00:57
Control Test		0.078	00:58
Air Blank		0.000	00:58
Diagnostics Check		OK	00:58

Data Download Date/Time: 04/30/2019 13:46

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CEPHUS,JAMIN,I
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: CEPHUS,JAMIN
ARRESTING OFFICER AGENCY: Perry PD
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: ADMIN
ARREST TIME: 15:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:30

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 04/18/2019 15:59:02

Diagnostics Check		OK	15:53
Air Blank		0.000	15:53
Control Test		0.080	15:53
Air Blank		0.000	15:54
Subject Sample	# 1	0.000	15:54
Breath Volume		3.101	15:54
Air Blank		0.000	15:55
Air Blank		0.000	15:57
Subject Sample	# 2	0.000	15:57
Breath Volume		3.167	15:57
Air Blank		0.000	15:58
Control Test		0.081	15:58
Air Blank		0.000	15:58
Diagnostics Check		OK	15:59

Data Download Date/Time: 04/30/2019 13:46

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARNOLD,JOHN,S
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: ARNOLD,JOHN
ARRESTING OFFICER AGENCY: Perry PD
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: ADMIN
ARREST TIME: 17:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:31

SUBJECT'S LAST NAME: THIGPEN

DATE/TIME OF TEST: 04/18/2019 17:57:54

Diagnostics Check		OK	17:51
Air Blank		0.000	17:52
Control Test		0.081	17:52
Air Blank		0.000	17:52
Subject Sample	# 1	0.000	17:53
Breath Volume		3.832	17:53
Air Blank		0.000	17:53
Air Blank		0.000	17:55
Subject Sample	# 2	0.000	17:56
Breath Volume		3.617	17:56
Air Blank		0.000	17:57
Control Test		0.080	17:57
Air Blank		0.000	17:57
Diagnostics Check		OK	17:57

Data Download Date/Time: 04/30/2019 13:46

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ERNST,WILLIAM,H
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: ERNST,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: MILES

DATE/TIME OF TEST: 04/26/2019 23:12:34

Diagnostics Check		OK	23:06
Air Blank		0.000	23:06
Control Test		0.080	23:06
Air Blank		0.000	23:07
Subject Sample	# 1	0.000	23:08
Breath Volume		3.371	23:08
Air Blank		0.000	23:08
Air Blank		0.000	23:10
Subject Sample	# 2	0.000	23:11
Breath Volume		3.140	23:11
Air Blank		0.000	23:11
Control Test		0.079	23:11
Air Blank		0.000	23:12
Diagnostics Check		OK	23:12

Data Download Date/Time: 04/30/2019 13:46

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE,RONALD,W
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LAMOUR,ZACH
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: OUK

DATE/TIME OF TEST: 04/11/2019 00:45:49

Diagnostics Check		OK	00:37	
Air Blank		0.000	00:37	
Control Test		0.078	00:38	
Air Blank		0.000	00:38	
Subject Sample	# 1		00:41	No Sample Provided
Breath Volume		0.000	00:41	
Air Blank		0.000	00:42	
Air Blank		0.000	00:44	
Subject Sample	# 2		00:44	Subject Test Refused
Breath Volume		0.000	00:44	
Air Blank		0.000	00:44	
Control Test		0.078	00:45	
Air Blank		0.000	00:45	
Diagnostics Check		OK	00:45	

Data Download Date/Time: 05/09/2019 09:56

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE,RONALD,W
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LAMOUR,ZACH
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: MALONE

DATE/TIME OF TEST: 04/12/2019 02:12:16

Diagnostics Check		OK	02:06	
Air Blank		0.000	02:06	
Control Test		0.079	02:07	
Air Blank		0.000	02:07	
Subject Sample	# 1	0.141	02:08	
Breath Volume		1.988	02:08	
Air Blank		0.000	02:08	
Air Blank		0.000	02:10	
Subject Sample	# 2		02:10	RFI Detect
Breath Volume		0.000	02:10	
Air Blank		0.000	02:11	
Control Test		0.077	02:11	
Air Blank		0.000	02:12	
Diagnostics Check		OK	02:12	

Data Download Date/Time: 05/09/2019 09:56

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE,RONALD,W
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LAMOUR,ZACH
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: MALONE

DATE/TIME OF TEST: 04/12/2019 02:14:12

Diagnostics Check	OK	02:13	
Air Blank		02:14	RFI Detect
Air Blank		02:14	RFI Detect

Data Download Date/Time: 05/09/2019 09:56

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE,RONALD,W
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LAMOUR,ZACH
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: MALONE

DATE/TIME OF TEST: 04/12/2019 02:22:13

Diagnostics Check		OK	02:15
Air Blank		0.000	02:16
Control Test		0.078	02:16
Air Blank		0.000	02:16
Subject Sample	# 1	0.139	02:17
Breath Volume		2.011	02:17
Air Blank		0.000	02:18
Air Blank		0.000	02:20
Subject Sample	# 2	0.137	02:20
Breath Volume		1.148	02:20
Air Blank		0.000	02:21
Control Test		0.077	02:21
Air Blank		0.000	02:22
Diagnostics Check		OK	02:22

Data Download Date/Time: 05/09/2019 09:56

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE,RONALD,W
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LAMOUR,ZACH
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: DROSIS

DATE/TIME OF TEST: 04/13/2019 01:26:50

Diagnostics Check		OK	01:15	
Air Blank		0.000	01:15	
Control Test		0.078	01:16	
Air Blank		0.000	01:16	
Subject Sample	# 1		01:19	No Sample Provided
Breath Volume		0.000	01:19	
Air Blank		0.000	01:20	
Air Blank		0.000	01:22	
Subject Sample	# 2		01:25	No Sample Provided
Breath Volume		0.000	01:25	
Air Blank		0.000	01:25	
Control Test		0.078	01:26	
Air Blank		0.000	01:26	
Diagnostics Check		OK	01:26	

Data Download Date/Time: 05/09/2019 09:56

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LESPERANCE,DANIEL,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCKENZIE,RONALD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: PHUNG

DATE/TIME OF TEST: 04/13/2019 03:10:43

Diagnostics Check		OK	02:59	
Air Blank		0.000	02:59	
Control Test		0.078	02:59	
Air Blank		0.000	03:00	
Subject Sample	# 1	0.100	03:03	Volume Not Met
Breath Volume		0.000	03:03	
Air Blank		0.000	03:04	
Air Blank		0.000	03:05	
Subject Sample	# 2	0.090	03:09	Volume Not Met
Breath Volume		0.398	03:09	
Air Blank		0.000	03:09	
Control Test		0.078	03:10	
Air Blank		0.000	03:10	
Diagnostics Check		OK	03:10	

Data Download Date/Time: 05/09/2019 09:56

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE,RONALD,W
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LESPERANCE,DANIEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:03

SUBJECT'S LAST NAME: AZHAR

DATE/TIME OF TEST: 04/13/2019 22:36:56

Diagnostics Check		OK	22:24	
Air Blank		0.000	22:24	
Control Test		0.078	22:25	
Air Blank		0.000	22:25	
Subject Sample	# 1		22:28	No Sample Provided
Breath Volume		0.000	22:28	
Air Blank		0.000	22:29	
Air Blank		0.000	22:31	
Subject Sample	# 2	0.272	22:32	
Breath Volume		1.906	22:32	
Air Blank		0.000	22:33	
Air Blank		0.000	22:34	
Subject Sample	# 3	0.269	22:35	
Breath Volume		1.609	22:35	
Air Blank		0.000	22:36	
Control Test		0.076	22:36	
Air Blank		0.000	22:36	
Diagnostics Check		OK	22:36	

Data Download Date/Time: 05/09/2019 09:56

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE,RONALD,W
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LAMOUR,ZACH
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 04/14/2019 02:49:48

Diagnostics Check		OK	02:43
Air Blank		0.000	02:43
Control Test		0.078	02:44
Air Blank		0.000	02:44
Subject Sample	# 1	0.162	02:45
Breath Volume		3.363	02:45
Air Blank		0.000	02:46
Air Blank		0.000	02:47
Subject Sample	# 2	0.157	02:48
Breath Volume		3.308	02:48
Air Blank		0.000	02:48
Control Test		0.076	02:49
Air Blank		0.000	02:49
Diagnostics Check		OK	02:49

Data Download Date/Time: 05/09/2019 09:57

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LESPERANCE,DANIEL,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LAMOUR,ZACHARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: BILANOVIC

DATE/TIME OF TEST: 04/19/2019 00:27:10

Diagnostics Check		OK	00:19
Air Blank		0.000	00:20
Control Test		0.077	00:20
Air Blank		0.000	00:21
Subject Sample	# 1	0.145	00:21
Breath Volume		1.148	00:21
Air Blank		0.000	00:22
Air Blank		0.000	00:24
Subject Sample	# 2	0.149	00:24
Breath Volume		1.562	00:24
Air Blank		0.000	00:25
Control Test		0.078	00:25
Air Blank		0.000	00:26
Diagnostics Check		OK	00:27

Data Download Date/Time: 05/09/2019 09:57

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVILA,NICOLE,M
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: AVILA,NICOLE
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 609878
DRY GAS EXPIRY DATE: 11/01/2019
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: FELSER

DATE/TIME OF TEST: 04/03/2019 02:18:08

Diagnostics Check		OK	02:12
Air Blank		0.000	02:12
Control Test		0.080	02:12
Air Blank		0.000	02:13
Subject Sample	# 1	0.146	02:13
Breath Volume		1.550	02:13
Air Blank		0.000	02:14
Air Blank		0.000	02:16
Subject Sample	# 2	0.146	02:16
Breath Volume		1.191	02:16
Air Blank		0.000	02:17
Control Test		0.080	02:17
Air Blank		0.000	02:18
Diagnostics Check		OK	02:18

Data Download Date/Time: 05/01/2019 00:08

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMBSH,CARLYTON,R
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: HAMBSH,CARLYTON
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 609878
DRY GAS EXPIRY DATE: 11/01/2019
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: SCHOFILL

DATE/TIME OF TEST: 04/05/2019 03:00:26

Diagnostics Check		OK	02:54
Air Blank		0.000	02:54
Control Test		0.080	02:54
Air Blank		0.000	02:55
Subject Sample	# 1	0.018	02:56
Breath Volume		2.867	02:56
Air Blank		0.000	02:56
Air Blank		0.000	02:58
Subject Sample	# 2	0.016	02:58
Breath Volume		1.617	02:58
Air Blank		0.000	02:59
Control Test		0.079	02:59
Air Blank		0.000	03:00
Diagnostics Check		OK	03:00

Data Download Date/Time: 05/01/2019 00:08

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDWARDS,JUSTIN,D
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: HICKS,Z
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 609878
DRY GAS EXPIRY DATE: 11/01/2019
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: KANYENGABU

DATE/TIME OF TEST: 04/13/2019 04:13:56

Diagnostics Check		OK	04:07
Air Blank		0.000	04:07
Control Test		0.080	04:07
Air Blank		0.000	04:08
Subject Sample	# 1	0.142	04:09
Breath Volume		2.695	04:09
Air Blank		0.000	04:10
Air Blank		0.000	04:11
Subject Sample	# 2	0.150	04:12
Breath Volume		2.296	04:12
Air Blank		0.000	04:13
Control Test		0.079	04:13
Air Blank		0.000	04:13
Diagnostics Check		OK	04:13

Data Download Date/Time: 05/01/2019 00:08

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDWARDS,JUSTIN,D
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: WILLIAMS,B
ARRESTING OFFICER AGENCY: Holly Hills PD
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 609878
DRY GAS EXPIRY DATE: 11/01/2019
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: RICHARDS

DATE/TIME OF TEST: 04/13/2019 22:12:17

Diagnostics Check		OK	22:00	
Air Blank		0.000	22:01	
Control Test		0.080	22:01	
Air Blank		0.000	22:02	
Subject Sample	# 1	0.158	22:05	Volume Not Met
Breath Volume		0.257	22:05	
Air Blank		0.000	22:05	
Air Blank		0.000	22:07	
Subject Sample	# 2	0.149	22:10	Volume Not Met
Breath Volume		0.234	22:10	
Air Blank		0.000	22:11	
Control Test		0.079	22:11	
Air Blank		0.000	22:12	
Diagnostics Check		OK	22:12	

Data Download Date/Time: 05/01/2019 00:08

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMBSH,CARLYTON,R
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: HAMBSH,CARLYTON
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 609878
DRY GAS EXPIRY DATE: 11/01/2019
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: BARKS

DATE/TIME OF TEST: 04/14/2019 00:28:59

Diagnostics Check		OK	00:23
Air Blank		0.000	00:23
Control Test		0.079	00:23
Air Blank		0.000	00:24
Subject Sample	# 1	0.207	00:24
Breath Volume		2.359	00:24
Air Blank		0.000	00:25
Air Blank		0.000	00:27
Subject Sample	# 2	0.204	00:27
Breath Volume		1.664	00:27
Air Blank		0.000	00:28
Control Test		0.079	00:28
Air Blank		0.000	00:28
Diagnostics Check		OK	00:28

Data Download Date/Time: 05/01/2019 00:08

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDWARDS,JUSTIN,D
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: DIAMOND,S
ARRESTING OFFICER AGENCY: South Daytona PD
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 609878
DRY GAS EXPIRY DATE: 11/01/2019
VIOLATION CODE: DUI
ARREST TIME: 08:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:02

SUBJECT'S LAST NAME: WALLER

DATE/TIME OF TEST: 04/27/2019 09:28:55

Diagnostics Check		OK	09:22
Air Blank		0.000	09:23
Control Test		0.079	09:23
Air Blank		0.000	09:24
Subject Sample	# 1	0.216	09:24
Breath Volume		3.335	09:24
Air Blank		0.000	09:25
Air Blank		0.000	09:26
Subject Sample	# 2	0.215	09:27
Breath Volume		3.097	09:27
Air Blank		0.000	09:28
Control Test		0.078	09:28
Air Blank		0.000	09:28
Diagnostics Check		OK	09:28

Data Download Date/Time: 05/01/2019 00:08

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STARLING,DARREN,J
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: STARLING,DARREN
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 609878
DRY GAS EXPIRY DATE: 11/01/2019
VIOLATION CODE: DUI
ARREST TIME: 18:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: SHELHORSE

DATE/TIME OF TEST: 04/27/2019 20:26:24

Diagnostics Check		OK	20:14	
Air Blank		0.000	20:15	
Control Test		0.080	20:15	
Air Blank		0.000	20:16	
Subject Sample	# 1	0.200	20:19	Volume Not Met
Breath Volume		1.085	20:19	
Air Blank		0.000	20:19	
Air Blank		0.000	20:21	
Subject Sample	# 2	0.191	20:24	Volume Not Met
Breath Volume		0.730	20:24	
Air Blank		0.000	20:25	
Control Test		0.079	20:25	
Air Blank		0.000	20:26	
Diagnostics Check		OK	20:26	

Data Download Date/Time: 05/01/2019 00:08

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STARLING,DARREN,J
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: STARLING,DARREN
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 609878
DRY GAS EXPIRY DATE: 11/01/2019
VIOLATION CODE: DUI
ARREST TIME: 18:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: SHELHORSE

DATE/TIME OF TEST: 04/27/2019 20:39:51

Diagnostics Check		OK	20:28	
Air Blank		0.000	20:28	
Control Test		0.078	20:29	
Air Blank		0.000	20:29	
Subject Sample	# 1	0.000	20:32	Volume Not Met
Breath Volume		0.000	20:32	
Air Blank		0.000	20:33	
Air Blank		0.000	20:35	
Subject Sample	# 2	0.000	20:38	Volume Not Met
Breath Volume		0.000	20:38	
Air Blank		0.000	20:38	
Control Test		0.078	20:39	
Air Blank		0.000	20:39	
Diagnostics Check		OK	20:39	

Data Download Date/Time: 05/01/2019 00:08

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STARLING,DARREN,J
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: STARLING,DARREN
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 609878
DRY GAS EXPIRY DATE: 11/01/2019
VIOLATION CODE: ADMIN
ARREST TIME: 18:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: STARLING

DATE/TIME OF TEST: 04/27/2019 20:55:20

Diagnostics Check		OK	20:49
Air Blank		0.000	20:49
Control Test		0.078	20:50
Air Blank		0.000	20:50
Subject Sample	# 1	0.000	20:51
Breath Volume		2.316	20:51
Air Blank		0.000	20:51
Air Blank		0.000	20:53
Subject Sample	# 2	0.000	20:53
Breath Volume		1.871	20:53
Air Blank		0.000	20:54
Control Test		0.078	20:54
Air Blank		0.000	20:55
Diagnostics Check		OK	20:55

Data Download Date/Time: 05/01/2019 00:08

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STARLING,DARREN,J
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: STARLING,DARREN
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 609878
DRY GAS EXPIRY DATE: 11/01/2019
VIOLATION CODE: DUI
ARREST TIME: 18:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: SHELHORSE

DATE/TIME OF TEST: 04/27/2019 21:03:03

Diagnostics Check		OK	20:57
Air Blank		0.000	20:57
Control Test		0.079	20:57
Air Blank		0.000	20:58
Subject Sample	# 1	0.209	20:58
Breath Volume		1.570	20:58
Air Blank		0.000	20:59
Air Blank		0.000	21:01
Subject Sample	# 2	0.215	21:01
Breath Volume		1.722	21:01
Air Blank		0.000	21:02
Control Test		0.079	21:02
Air Blank		0.000	21:02
Diagnostics Check		OK	21:03

Data Download Date/Time: 05/01/2019 00:09

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KURTZ,THEODORE,F
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: KURTZ,THEODORE
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 609878
DRY GAS EXPIRY DATE: 11/01/2019
VIOLATION CODE: DUI
ARREST TIME: 20:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: ROYER

DATE/TIME OF TEST: 04/27/2019 21:38:27

Diagnostics Check		OK	21:30
Air Blank		0.000	21:30
Control Test		0.078	21:30
Air Blank		0.000	21:31
Subject Sample	# 1	0.181	21:31
Breath Volume		2.105	21:31
Air Blank		0.000	21:32
Air Blank		0.000	21:34
Subject Sample	# 2	0.178	21:36
Breath Volume		1.800	21:36
Air Blank		0.000	21:37
Control Test		0.079	21:37
Air Blank		0.000	21:38
Diagnostics Check		OK	21:38

Data Download Date/Time: 05/01/2019 00:09

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUFFY,J,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: JEAN-BAPTISTE,L
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: WOOD

DATE/TIME OF TEST: 04/01/2019 04:56:08

Diagnostics Check		OK	04:49
Air Blank		0.000	04:50
Control Test		0.081	04:50
Air Blank		0.000	04:50
Subject Sample	# 1	0.079	04:51
Breath Volume		2.601	04:51
Air Blank		0.000	04:52
Air Blank		0.000	04:53
Subject Sample	# 2	0.076	04:54
Breath Volume		2.691	04:54
Air Blank		0.000	04:55
Control Test		0.082	04:55
Air Blank		0.000	04:56
Diagnostics Check		OK	04:56

Data Download Date/Time: 04/17/2019 14:44

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCBRIDE,TIMOTHY,A
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: MCBRIDE,TIMOTHY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:59

SUBJECT'S LAST NAME: GONZALEZ TORRES

DATE/TIME OF TEST: 04/03/2019 05:28:05

Diagnostics Check		OK	05:21
Air Blank		0.000	05:21
Control Test		0.082	05:21
Air Blank		0.000	05:22
Subject Sample	# 1	0.138	05:22
Breath Volume		1.816	05:22
Air Blank		0.000	05:23
Air Blank		0.000	05:25
Subject Sample	# 2	0.149	05:26
Breath Volume		2.785	05:26
Air Blank		0.000	05:27
Control Test		0.081	05:27
Air Blank		0.000	05:28
Diagnostics Check		OK	05:28

Data Download Date/Time: 04/17/2019 14:44

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCBRIDE,TIMOTHY,A
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: MCBRIDE,TIMOTHY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: GONZALEZ SILVA

DATE/TIME OF TEST: 04/06/2019 04:30:33

Diagnostics Check		OK	04:23
Air Blank		0.000	04:24
Control Test		0.082	04:24
Air Blank		0.000	04:25
Subject Sample	# 1	0.120	04:25
Breath Volume		3.695	04:25
Air Blank		0.000	04:26
Air Blank		0.000	04:28
Subject Sample	# 2	0.118	04:28
Breath Volume		3.710	04:28
Air Blank		0.000	04:29
Control Test		0.080	04:29
Air Blank		0.000	04:30
Diagnostics Check		OK	04:30

Data Download Date/Time: 04/17/2019 14:44

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,JAMES,R
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: TELUS,JEAN
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: PALACIOS

DATE/TIME OF TEST: 04/07/2019 04:07:18

Diagnostics Check		OK	04:00
Air Blank		0.000	04:01
Control Test		0.082	04:01
Air Blank		0.000	04:02
Subject Sample	# 1	0.168	04:02
Breath Volume		2.035	04:02
Air Blank		0.000	04:03
Air Blank		0.000	04:05
Subject Sample	# 2	0.164	04:05
Breath Volume		1.828	04:05
Air Blank		0.000	04:06
Control Test		0.080	04:06
Air Blank		0.000	04:07
Diagnostics Check		OK	04:07

Data Download Date/Time: 04/17/2019 14:44

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROMAN,A,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: OLEARY,P
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: TAMBURRI

DATE/TIME OF TEST: 04/07/2019 04:56:36

Diagnostics Check		OK	04:50
Air Blank		0.000	04:50
Control Test		0.082	04:50
Air Blank		0.000	04:51
Subject Sample	# 1	0.213	04:51
Breath Volume		2.144	04:51
Air Blank		0.000	04:52
Air Blank		0.000	04:54
Subject Sample	# 2	0.216	04:54
Breath Volume		2.921	04:54
Air Blank		0.000	04:55
Control Test		0.081	04:55
Air Blank		0.000	04:56
Diagnostics Check		OK	04:56

Data Download Date/Time: 04/17/2019 14:44

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,JAMES,R
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: RIDGEWAY,CAITLIN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: MUNIZ VIERA

DATE/TIME OF TEST: 04/07/2019 05:17:16

Diagnostics Check		OK	05:13	
Air Blank		0.000	05:13	
Control Test		0.081	05:14	
Air Blank		0.000	05:14	
Subject Sample	# 1		05:15	Subject Test Refused
Breath Volume		0.000	05:15	
Air Blank		0.000	05:16	
Control Test		0.082	05:16	
Air Blank			05:16	Sequence Aborted
Air Blank		0.000	05:17	

Data Download Date/Time: 04/17/2019 14:44

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,CHRISTINE,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: HALL,CHRISTINE
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: CRISTOBAL

DATE/TIME OF TEST: 04/07/2019 19:50:04

Diagnostics Check		OK	19:43
Air Blank		0.000	19:43
Control Test		0.082	19:44
Air Blank		0.000	19:44
Subject Sample	# 1	0.195	19:45
Breath Volume		2.828	19:45
Air Blank		0.000	19:46
Air Blank		0.000	19:47
Subject Sample	# 2	0.198	19:48
Breath Volume		3.109	19:48
Air Blank		0.000	19:49
Control Test		0.081	19:49
Air Blank		0.000	19:49
Diagnostics Check		OK	19:50

Data Download Date/Time: 04/17/2019 14:44

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCBRIDE,TIMOTHY,A
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: MARTINS,K
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: HODSON

DATE/TIME OF TEST: 04/08/2019 02:20:03

Diagnostics Check		OK	02:13
Air Blank		0.000	02:14
Control Test		0.082	02:14
Air Blank		0.000	02:15
Subject Sample	# 1	0.015	02:15
Breath Volume		1.613	02:15
Air Blank		0.000	02:16
Air Blank		0.000	02:17
Subject Sample	# 2	0.016	02:18
Breath Volume		1.679	02:18
Air Blank		0.000	02:19
Control Test		0.082	02:19
Air Blank		0.000	02:19
Diagnostics Check		OK	02:20

Data Download Date/Time: 04/17/2019 14:44

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCBRIDE,TIMOTHY,A
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: MCBRIDE,TIMOTHY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 04/11/2019 22:50:35

Diagnostics Check		OK	22:43
Air Blank		0.000	22:44
Control Test		0.083	22:44
Air Blank		0.000	22:44
Subject Sample	# 1	0.126	22:45
Breath Volume		3.257	22:45
Air Blank		0.000	22:46
Air Blank		0.000	22:48
Subject Sample	# 2	0.127	22:48
Breath Volume		3.761	22:48
Air Blank		0.000	22:49
Control Test		0.083	22:49
Air Blank		0.000	22:50
Diagnostics Check		OK	22:50

Data Download Date/Time: 04/17/2019 14:44

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,EARTHEN,A
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: BROWN,EARTHEN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 04/12/2019 23:25:33

Diagnostics Check		OK	23:17
Air Blank		0.000	23:17
Control Test		0.083	23:17
Air Blank		0.000	23:18
Subject Sample	# 1	0.178	23:20
Breath Volume		1.175	23:20
Air Blank		0.000	23:21
Air Blank		0.000	23:23
Subject Sample	# 2	0.166	23:23
Breath Volume		1.296	23:23
Air Blank		0.000	23:24
Control Test		0.079	23:24
Air Blank		0.000	23:25
Diagnostics Check		OK	23:25

Data Download Date/Time: 04/17/2019 14:44

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCBRIDE,TIMOTHY,A
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: MCBRIDE,TIMOTHY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 04/16/2019 00:22:14

Diagnostics Check		OK	00:12
Air Blank		0.000	00:13
Control Test		0.084	00:13
Air Blank		0.000	00:14
Subject Sample	# 1	0.140	00:16
Breath Volume		1.250	00:16
Air Blank		0.000	00:17
Air Blank		0.000	00:18
Subject Sample	# 2	0.135	00:20
Breath Volume		1.199	00:20
Air Blank		0.000	00:21
Control Test		0.081	00:21
Air Blank		0.000	00:22
Diagnostics Check		OK	00:22

Data Download Date/Time: 04/17/2019 14:44

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOTELLO,A,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: BOTELLO,A
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: MARINES

DATE/TIME OF TEST: 04/16/2019 01:41:29

Diagnostics Check		OK	01:34
Air Blank		0.000	01:35
Control Test		0.081	01:35
Air Blank		0.000	01:35
Subject Sample	# 1	0.172	01:36
Breath Volume		2.734	01:36
Air Blank		0.000	01:37
Air Blank		0.000	01:39
Subject Sample	# 2	0.175	01:39
Breath Volume		2.324	01:39
Air Blank		0.000	01:40
Control Test		0.082	01:40
Air Blank		0.000	01:41
Diagnostics Check		OK	01:41

Data Download Date/Time: 04/17/2019 14:44

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCBRIDE,TIMOTHY,A
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: MCBRIDE,TIMOTHY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: HERBERT

DATE/TIME OF TEST: 04/17/2019 01:50:28

Diagnostics Check		OK	01:41
Air Blank		0.000	01:42
Control Test		0.082	01:42
Air Blank		0.000	01:43
Subject Sample	# 1	0.178	01:45
Breath Volume		1.921	01:45
Air Blank		0.000	01:46
Air Blank		0.000	01:48
Subject Sample	# 2	0.171	01:48
Breath Volume		2.183	01:48
Air Blank		0.000	01:49
Control Test		0.080	01:49
Air Blank		0.000	01:50
Diagnostics Check		OK	01:50

Data Download Date/Time: 04/17/2019 14:44

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDIN,JEREMY,L
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: TOMYN,APRIL
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/17/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 06:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:21

SUBJECT'S LAST NAME: CROSS

DATE/TIME OF TEST: 04/18/2019 07:48:31

Diagnostics Check		OK	07:42
Air Blank		0.000	07:42
Control Test		0.083	07:42
Air Blank		0.000	07:43
Subject Sample	# 1	0.210	07:43
Breath Volume		1.304	07:43
Air Blank		0.000	07:44
Air Blank		0.000	07:46
Subject Sample	# 2	0.204	07:46
Breath Volume		1.726	07:46
Air Blank		0.000	07:47
Control Test		0.080	07:47
Air Blank		0.000	07:48
Diagnostics Check		OK	07:48

Data Download Date/Time: 05/08/2019 13:05

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JOSIAH,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: WINCELOWICZ,LARRY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/17/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: LOPEZ DOMINGO

DATE/TIME OF TEST: 04/20/2019 04:20:30

Diagnostics Check		OK	04:11
Air Blank		0.000	04:12
Control Test		0.083	04:12
Air Blank		0.000	04:13
Subject Sample	# 1	0.181	04:15
Breath Volume		2.835	04:15
Air Blank		0.000	04:16
Air Blank		0.000	04:18
Subject Sample	# 2	0.180	04:18
Breath Volume		1.195	04:18
Air Blank		0.000	04:19
Control Test		0.080	04:19
Air Blank		0.000	04:20
Diagnostics Check		OK	04:20

Data Download Date/Time: 05/08/2019 13:05

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOULD,LEONARD,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: BRAY,CADE
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/17/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: ROSENTHAL

DATE/TIME OF TEST: 04/23/2019 23:34:47

Diagnostics Check		OK	23:31	
Air Blank		0.000	23:32	
Control Test		0.083	23:32	
Air Blank		0.000	23:33	
Subject Sample	# 1		23:33	Subject Test Refused
Breath Volume		0.000	23:33	
Air Blank		0.000	23:33	
Control Test		0.083	23:34	
Air Blank		0.000	23:34	
Diagnostics Check		OK	23:34	

Data Download Date/Time: 05/08/2019 13:05

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAZURYK,MICHAEL,S
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: MAZURYK,MICHAEL
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 04/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:54

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 04/17/2019 22:17:41

Diagnostics Check		OK	22:14	
Air Blank		0.000	22:15	
Control Test		0.081	22:15	
Air Blank		0.000	22:15	
Subject Sample	# 1		22:16	Subject Test Refused
Breath Volume		0.000	22:16	
Air Blank		0.000	22:16	
Control Test		0.079	22:17	
Air Blank		0.000	22:17	
Diagnostics Check		OK	22:17	

Data Download Date/Time: 05/07/2019 13:27

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAETSCH,JOSH,E
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PAETSCH,JOSH
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 04/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:46

SUBJECT'S LAST NAME: MYERS

DATE/TIME OF TEST: 04/20/2019 21:38:50

Diagnostics Check		OK	21:27
Air Blank		0.000	21:28
Control Test		0.081	21:28
Air Blank		0.000	21:28
Subject Sample	# 1	0.266	21:31
Breath Volume		2.773	21:31
Air Blank		0.000	21:32
Air Blank		0.000	21:33
Subject Sample	# 2	0.227	21:34
Breath Volume		1.468	21:34
Air Blank		0.000	21:34
Air Blank		0.000	21:36
Subject Sample	# 3	0.244	21:37
Breath Volume		2.117	21:37
Air Blank		0.000	21:37
Control Test		0.079	21:38
Air Blank		0.000	21:38
Diagnostics Check		OK	21:38

Data Download Date/Time: 05/07/2019 13:27

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: PACHECO,GEORGE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: ALANIZ

DATE/TIME OF TEST: 04/05/2019 03:54:27

Diagnostics Check		OK	03:51	
Air Blank		0.000	03:52	
Control Test		0.078	03:52	
Air Blank		0.000	03:52	
Subject Sample	# 1		03:53	Subject Test Refused
Breath Volume		0.000	03:53	
Air Blank		0.000	03:53	
Control Test		0.077	03:53	
Air Blank		0.000	03:54	
Diagnostics Check		OK	03:54	

Data Download Date/Time: 04/11/2019 09:44

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: VARGAS,DAEL
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: PAEZ

DATE/TIME OF TEST: 04/06/2019 02:09:22

Diagnostics Check		OK	02:06	
Air Blank		0.000	02:06	
Control Test		0.078	02:07	
Air Blank		0.000	02:07	
Subject Sample	# 1		02:07	Subject Test Refused
Breath Volume		0.000	02:07	
Air Blank		0.000	02:08	
Control Test		0.078	02:08	
Air Blank		0.000	02:09	
Diagnostics Check		OK	02:09	

Data Download Date/Time: 04/11/2019 09:44

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: HERRERO-LAZO

DATE/TIME OF TEST: 04/06/2019 04:32:27

Diagnostics Check		OK	04:25
Air Blank		0.000	04:26
Control Test		0.078	04:26
Air Blank		0.000	04:26
Subject Sample	# 1	0.041	04:28
Breath Volume		1.683	04:28
Air Blank		0.000	04:28
Air Blank		0.000	04:30
Subject Sample	# 2	0.041	04:30
Breath Volume		1.531	04:30
Air Blank		0.000	04:31
Control Test		0.078	04:31
Air Blank		0.000	04:32
Diagnostics Check		OK	04:32

Data Download Date/Time: 04/11/2019 09:44

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: QUINTANA,JORGE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: DE AHORA

DATE/TIME OF TEST: 04/07/2019 00:46:58

Diagnostics Check		OK	00:40
Air Blank		0.000	00:40
Control Test		0.078	00:40
Air Blank		0.000	00:41
Subject Sample	# 1	0.198	00:42
Breath Volume		1.488	00:42
Air Blank		0.000	00:42
Air Blank		0.000	00:44
Subject Sample	# 2	0.200	00:45
Breath Volume		2.097	00:45
Air Blank		0.000	00:46
Control Test		0.077	00:46
Air Blank		0.000	00:46
Diagnostics Check		OK	00:46

Data Download Date/Time: 04/11/2019 09:44

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: BIANCHI,MARIO
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: CASTRILLON

DATE/TIME OF TEST: 04/07/2019 02:34:10

Diagnostics Check		OK	02:27
Air Blank		0.000	02:28
Control Test		0.079	02:28
Air Blank		0.000	02:29
Subject Sample	# 1	0.128	02:29
Breath Volume		1.632	02:29
Air Blank		0.000	02:30
Air Blank		0.000	02:31
Subject Sample	# 2	0.126	02:32
Breath Volume		1.652	02:32
Air Blank		0.000	02:33
Control Test		0.077	02:33
Air Blank		0.000	02:34
Diagnostics Check		OK	02:34

Data Download Date/Time: 04/11/2019 09:44

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 05:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:41

SUBJECT'S LAST NAME: SALINAS

DATE/TIME OF TEST: 04/07/2019 06:14:23

Diagnostics Check		OK	06:04
Air Blank		0.000	06:05
Control Test		0.078	06:05
Air Blank		0.000	06:06
Subject Sample	# 1	0.171	06:06
Breath Volume		2.148	06:06
Air Blank		0.000	06:07
Air Blank		0.000	06:09
Subject Sample	# 2	0.149	06:09
Breath Volume		1.449	06:09
Air Blank		0.000	06:10
Air Blank		0.000	06:12
Subject Sample	# 3	0.163	06:12
Breath Volume		2.554	06:12
Air Blank		0.000	06:13
Control Test		0.078	06:13
Air Blank		0.000	06:14
Diagnostics Check		OK	06:14

Data Download Date/Time: 04/11/2019 09:44

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: SLIMAK,MARK
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: REYNOLDS

DATE/TIME OF TEST: 04/11/2019 02:00:18

Diagnostics Check		OK	01:57	
Air Blank		0.000	01:57	
Control Test		0.079	01:58	
Air Blank		0.000	01:58	
Subject Sample	# 1		01:58	Subject Test Refused
Breath Volume		0.000	01:58	
Air Blank		0.000	01:59	
Control Test		0.078	01:59	
Air Blank		0.000	02:00	
Diagnostics Check		OK	02:00	

Data Download Date/Time: 04/11/2019 09:44

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: QUINTANA,JORGE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: FERRIN FUNDORA

DATE/TIME OF TEST: 04/11/2019 03:59:45

Diagnostics Check		OK	03:51
Air Blank		0.000	03:52
Control Test		0.079	03:52
Air Blank		0.000	03:53
Subject Sample	# 1	0.117	03:54
Breath Volume		1.367	03:54
Air Blank		0.000	03:55
Air Blank		0.000	03:57
Subject Sample	# 2	0.119	03:58
Breath Volume		1.261	03:58
Air Blank		0.000	03:58
Control Test		0.078	03:59
Air Blank		0.000	03:59
Diagnostics Check		OK	03:59

Data Download Date/Time: 04/11/2019 09:44

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 04/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:03

SUBJECT'S LAST NAME: RUIZ

DATE/TIME OF TEST: 04/12/2019 05:30:20

Diagnostics Check		OK	05:27	
Air Blank		0.000	05:27	
Control Test		0.078	05:28	
Air Blank		0.000	05:28	
Subject Sample	# 1		05:28	Subject Test Refused
Breath Volume		0.000	05:28	
Air Blank		0.000	05:29	
Control Test		0.078	05:29	
Air Blank		0.000	05:30	
Diagnostics Check		OK	05:30	

Data Download Date/Time: 05/02/2019 10:20

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: VARGAS,DAEL
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 04/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: CABEZAS

DATE/TIME OF TEST: 04/13/2019 01:25:18

Diagnostics Check		OK	01:22	
Air Blank		0.000	01:22	
Control Test		0.078	01:23	
Air Blank		0.000	01:23	
Subject Sample	# 1		01:23	Subject Test Refused
Breath Volume		0.000	01:23	
Air Blank		0.000	01:24	
Control Test		0.078	01:24	
Air Blank		0.000	01:25	
Diagnostics Check		OK	01:25	

Data Download Date/Time: 05/02/2019 10:20

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: QUINTANA,JORGE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 04/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: SOLIS

DATE/TIME OF TEST: 04/13/2019 03:31:38

Diagnostics Check		OK	03:25
Air Blank		0.000	03:25
Control Test		0.078	03:26
Air Blank		0.000	03:26
Subject Sample	# 1	0.150	03:27
Breath Volume		2.507	03:27
Air Blank		0.000	03:27
Air Blank		0.000	03:29
Subject Sample	# 2	0.151	03:30
Breath Volume		2.527	03:30
Air Blank		0.000	03:30
Control Test		0.077	03:31
Air Blank		0.000	03:31
Diagnostics Check		OK	03:31

Data Download Date/Time: 05/02/2019 10:20

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 04/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:02

SUBJECT'S LAST NAME: GRANIZO-SUAREZ

DATE/TIME OF TEST: 04/13/2019 05:32:23

Diagnostics Check		OK	05:25
Air Blank		0.000	05:26
Control Test		0.077	05:26
Air Blank		0.000	05:27
Subject Sample	# 1	0.209	05:27
Breath Volume		1.558	05:27
Air Blank		0.000	05:28
Air Blank		0.000	05:29
Subject Sample	# 2	0.208	05:30
Breath Volume		1.566	05:30
Air Blank		0.000	05:31
Control Test		0.077	05:31
Air Blank		0.000	05:32
Diagnostics Check		OK	05:32

Data Download Date/Time: 05/02/2019 10:20

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: VARGAS,DAEL
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 04/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 04/17/2019 02:40:12

Diagnostics Check		OK	02:37	
Air Blank		0.000	02:37	
Control Test		0.079	02:38	
Air Blank		0.000	02:38	
Subject Sample	# 1		02:38	Subject Test Refused
Breath Volume		0.000	02:38	
Air Blank		0.000	02:39	
Control Test		0.078	02:39	
Air Blank		0.000	02:40	
Diagnostics Check		OK	02:40	

Data Download Date/Time: 05/02/2019 10:20

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 04/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: CRIOLLO

DATE/TIME OF TEST: 04/18/2019 01:45:58

Diagnostics Check		OK	01:43	
Air Blank		0.000	01:43	
Control Test		0.078	01:43	
Air Blank		0.000	01:44	
Subject Sample	# 1		01:44	Subject Test Refused
Breath Volume		0.000	01:44	
Air Blank		0.000	01:45	
Control Test		0.078	01:45	
Air Blank		0.000	01:45	
Diagnostics Check		OK	01:45	

Data Download Date/Time: 05/02/2019 10:20

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: QUINTANA,JORGE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 04/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: MARQUEZ

DATE/TIME OF TEST: 04/19/2019 04:40:14

Diagnostics Check		OK	04:37	
Air Blank		0.000	04:37	
Control Test		0.078	04:38	
Air Blank		0.000	04:38	
Subject Sample	# 1		04:38	Subject Test Refused
Breath Volume		0.000	04:38	
Air Blank		0.000	04:39	
Control Test		0.077	04:39	
Air Blank		0.000	04:40	
Diagnostics Check		OK	04:40	

Data Download Date/Time: 05/02/2019 10:20

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 04/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:47

SUBJECT'S LAST NAME: YU

DATE/TIME OF TEST: 04/19/2019 05:15:52

Diagnostics Check		OK	05:09
Air Blank		0.000	05:09
Control Test		0.077	05:09
Air Blank		0.000	05:10
Subject Sample	# 1	0.112	05:11
Breath Volume		1.429	05:11
Air Blank		0.000	05:12
Air Blank		0.000	05:13
Subject Sample	# 2	0.111	05:14
Breath Volume		1.742	05:14
Air Blank		0.000	05:14
Control Test		0.078	05:15
Air Blank		0.000	05:15
Diagnostics Check		OK	05:15

Data Download Date/Time: 05/02/2019 10:20

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 04/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: SANDIGO-SEQUEIRA

DATE/TIME OF TEST: 04/21/2019 04:42:51

Diagnostics Check		OK	04:34
Air Blank		0.000	04:34
Control Test		0.080	04:34
Air Blank		0.000	04:35
Subject Sample	# 1	0.204	04:37
Breath Volume		1.503	04:37
Air Blank		0.000	04:37
Air Blank		0.000	04:39
Subject Sample	# 2	0.214	04:41
Breath Volume		1.828	04:41
Air Blank		0.000	04:41
Control Test		0.078	04:42
Air Blank		0.000	04:42
Diagnostics Check		OK	04:42

Data Download Date/Time: 05/02/2019 10:20

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 04/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 05:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:29

SUBJECT'S LAST NAME: PONCE DE LEON

DATE/TIME OF TEST: 04/21/2019 05:59:09

Diagnostics Check		OK	05:53
Air Blank		0.000	05:53
Control Test		0.080	05:53
Air Blank		0.000	05:54
Subject Sample	# 1	0.000	05:54
Breath Volume		1.847	05:54
Air Blank		0.000	05:55
Air Blank		0.000	05:57
Subject Sample	# 2	0.000	05:57
Breath Volume		1.910	05:57
Air Blank		0.000	05:58
Control Test		0.079	05:58
Air Blank		0.000	05:59
Diagnostics Check		OK	05:59

Data Download Date/Time: 05/02/2019 10:20

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 04/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: MARTINEZ-SOSSA

DATE/TIME OF TEST: 04/24/2019 03:12:46

Diagnostics Check		OK	03:06
Air Blank		0.000	03:06
Control Test		0.079	03:07
Air Blank		0.000	03:07
Subject Sample	# 1	0.238	03:08
Breath Volume		2.367	03:08
Air Blank		0.000	03:08
Air Blank		0.000	03:10
Subject Sample	# 2	0.237	03:11
Breath Volume		2.070	03:11
Air Blank		0.000	03:11
Control Test		0.078	03:12
Air Blank		0.000	03:12
Diagnostics Check		OK	03:12

Data Download Date/Time: 05/02/2019 10:20

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 04/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: BUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: CARRILLO

DATE/TIME OF TEST: 04/29/2019 01:54:57

Diagnostics Check		OK	01:52	
Air Blank		0.000	01:52	
Control Test		0.078	01:52	
Air Blank		0.000	01:53	
Subject Sample	# 1		01:53	Subject Test Refused
Breath Volume		0.000	01:53	
Air Blank		0.000	01:54	
Control Test		0.078	01:54	
Air Blank		0.000	01:54	
Diagnostics Check		OK	01:54	

Data Download Date/Time: 05/02/2019 10:20

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAVAS,DAVID,E
BREATH TEST OPERATOR AGENCY: BAL HARBOUR VIL PD
ARRESTING OFFICER: BLANCO,ERICA
ARRESTING OFFICER AGENCY: BAL HARBOUR VIL PD
LAST AGENCY INSPECTION: 04/15/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: MAXIMOVA

DATE/TIME OF TEST: 04/22/2019 04:03:39

Diagnostics Check		OK	03:59	
Air Blank		0.000	03:59	
Control Test		0.079	03:59	
Air Blank		0.000	04:00	
Subject Sample	# 1		04:02	Subject Test Refused
Breath Volume		0.000	04:02	
Air Blank			04:02	Purge Fail
Air Blank			04:03	Purge Fail

Data Download Date/Time: 05/08/2019 11:03

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IINO,RYUNOSUKE,
BREATH TEST OPERATOR AGENCY: AVENTURA PD
ARRESTING OFFICER: STEPHENSON,TIFANY
ARRESTING OFFICER AGENCY: SUNNY ISLES BEACH PD
LAST AGENCY INSPECTION: 04/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: STAN

DATE/TIME OF TEST: 04/26/2019 21:47:01

Diagnostics Check		OK	21:37
Air Blank		0.000	21:37
Control Test		0.078	21:37
Air Blank		0.000	21:38
Subject Sample	# 1	0.302	21:39
Breath Volume		2.046	21:39
Air Blank		0.000	21:39
Air Blank		0.000	21:41
Subject Sample	# 2	0.277	21:42
Breath Volume		1.156	21:42
Air Blank		0.000	21:43
Air Blank		0.000	21:44
Subject Sample	# 3	0.273	21:45
Breath Volume		2.796	21:45
Air Blank		0.000	21:46
Control Test		0.077	21:46
Air Blank		0.000	21:46
Diagnostics Check		OK	21:47

Data Download Date/Time: 05/08/2019 11:03

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AJO,MICHAEL,D
BREATH TEST OPERATOR AGENCY: NORTH MIAMI BEACH PD
ARRESTING OFFICER: AJO,MICHAEL
ARRESTING OFFICER AGENCY: NORTH MIAMI BEACH PD
LAST AGENCY INSPECTION: 03/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: CALIXTE

DATE/TIME OF TEST: 04/03/2019 03:10:19

Diagnostics Check		OK	03:03
Air Blank		0.000	03:04
Control Test		0.079	03:04
Air Blank		0.000	03:05
Subject Sample	# 1	0.045	03:05
Breath Volume		1.820	03:05
Air Blank		0.000	03:06
Air Blank		0.000	03:08
Subject Sample	# 2	0.043	03:08
Breath Volume		1.105	03:08
Air Blank		0.000	03:09
Control Test		0.078	03:09
Air Blank		0.000	03:10
Diagnostics Check		OK	03:10

Data Download Date/Time: 04/25/2019 13:27

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGRATH,ROBERT,E
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: MCGRATH,ROBERT
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 03/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: LAWRENCE

DATE/TIME OF TEST: 04/09/2019 20:44:46

Diagnostics Check		OK	20:32	
Air Blank		0.000	20:33	
Control Test		0.077	20:33	
Air Blank		0.000	20:34	
Subject Sample	# 1	0.235	20:37	Volume Not Met
Breath Volume		0.023	20:37	
Air Blank		0.000	20:38	
Air Blank		0.000	20:39	
Subject Sample	# 2	0.266	20:42	Volume Not Met
Breath Volume		0.015	20:42	
Air Blank		0.000	20:43	
Control Test		0.075	20:44	
Air Blank		0.000	20:44	
Diagnostics Check		OK	20:44	

Data Download Date/Time: 04/25/2019 13:27

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGRATH,ROBERT,E
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: MCGRATH,ROBERT
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 03/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: LAWRENCE

DATE/TIME OF TEST: 04/09/2019 21:45:17

Diagnostics Check		OK	21:42	
Air Blank		0.000	21:42	
Control Test		0.078	21:43	
Air Blank		0.000	21:43	
Subject Sample	# 1		21:43	Subject Test Refused
Breath Volume		0.000	21:43	
Air Blank		0.000	21:44	
Control Test		0.079	21:44	
Air Blank		0.000	21:45	
Diagnostics Check		OK	21:45	

Data Download Date/Time: 04/25/2019 13:27

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PACHECO,GEORGE,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: PACHECO,GEORGE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 03/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 05:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:05

SUBJECT'S LAST NAME: CHARLES

DATE/TIME OF TEST: 04/10/2019 06:33:09

Diagnostics Check		OK	06:26
Air Blank		0.000	06:27
Control Test		0.079	06:27
Air Blank		0.000	06:28
Subject Sample	# 1	0.100	06:28
Breath Volume		1.886	06:28
Air Blank		0.000	06:29
Air Blank		0.000	06:31
Subject Sample	# 2	0.099	06:31
Breath Volume		1.437	06:31
Air Blank		0.000	06:32
Control Test		0.079	06:32
Air Blank		0.000	06:33
Diagnostics Check		OK	06:33

Data Download Date/Time: 04/25/2019 13:27

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KONETZ,JONATHAN,D
BREATH TEST OPERATOR AGENCY: SUNNY ISLES BEACH PD
ARRESTING OFFICER: KONETZ,JONATHAN
ARRESTING OFFICER AGENCY: SUNNY ISLES BEACH PD
LAST AGENCY INSPECTION: 03/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: PALAMARCHUK

DATE/TIME OF TEST: 04/13/2019 00:02:41

Diagnostics Check		OK	23:59	
Air Blank		0.000	00:00	
Control Test		0.079	00:00	
Air Blank		0.000	00:01	
Subject Sample	# 1		00:01	Subject Test Refused
Breath Volume		0.000	00:01	
Air Blank		0.000	00:01	
Control Test		0.079	00:02	
Air Blank		0.000	00:02	
Diagnostics Check		OK	00:02	

Data Download Date/Time: 04/25/2019 13:27

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GABRIEL,EVENS,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: RODRIGUEZ,YURISLAIDIS
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 03/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 04/13/2019 19:05:12

Diagnostics Check		OK	19:02	
Air Blank		0.000	19:02	
Control Test		0.078	19:03	
Air Blank		0.000	19:03	
Subject Sample	# 1		19:03	Subject Test Refused
Breath Volume		0.000	19:03	
Air Blank		0.000	19:04	
Control Test		0.079	19:04	
Air Blank		0.000	19:05	
Diagnostics Check		OK	19:05	

Data Download Date/Time: 04/25/2019 13:27

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGRATH,ROBERT,E
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: MCGRATH,ROBERT
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 03/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 04/15/2019 03:45:50

Diagnostics Check		OK	03:37
Air Blank		0.000	03:38
Control Test		0.080	03:38
Air Blank		0.000	03:38
Subject Sample	# 1	0.149	03:40
Breath Volume		1.218	03:40
Air Blank		0.000	03:41
Air Blank		0.000	03:43
Subject Sample	# 2	0.149	03:44
Breath Volume		1.496	03:44
Air Blank		0.000	03:44
Control Test		0.078	03:45
Air Blank		0.000	03:45
Diagnostics Check		OK	03:45

Data Download Date/Time: 04/25/2019 13:27

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: LILJENQUIST

DATE/TIME OF TEST: 04/05/2019 04:35:33

Diagnostics Check		OK	04:29
Air Blank		0.000	04:29
Control Test		0.077	04:30
Air Blank		0.000	04:30
Subject Sample	# 1	0.192	04:31
Breath Volume		3.421	04:31
Air Blank		0.000	04:31
Air Blank		0.000	04:33
Subject Sample	# 2	0.194	04:33
Breath Volume		3.742	04:33
Air Blank		0.000	04:34
Control Test		0.077	04:34
Air Blank		0.000	04:35
Diagnostics Check		OK	04:35

Data Download Date/Time: 04/11/2019 09:49

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 04/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: CARBALLO

DATE/TIME OF TEST: 04/26/2019 03:33:04

Diagnostics Check		OK	03:26
Air Blank		0.000	03:27
Control Test		0.078	03:27
Air Blank		0.000	03:28
Subject Sample	# 1	0.220	03:28
Breath Volume		2.390	03:28
Air Blank		0.000	03:29
Air Blank		0.000	03:31
Subject Sample	# 2	0.219	03:31
Breath Volume		2.046	03:31
Air Blank		0.000	03:32
Control Test		0.078	03:32
Air Blank		0.000	03:32
Diagnostics Check		OK	03:33

Data Download Date/Time: 05/02/2019 10:18

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLAFLIN,DAVID,S
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GOULD,LENNY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: SIDWELL

DATE/TIME OF TEST: 04/05/2019 22:56:33

Diagnostics Check		OK	22:50
Air Blank		0.000	22:50
Control Test		0.080	22:50
Air Blank		0.000	22:51
Subject Sample	# 1	0.037	22:51
Breath Volume		2.683	22:51
Air Blank		0.000	22:52
Air Blank		0.000	22:54
Subject Sample	# 2	0.040	22:54
Breath Volume		2.617	22:54
Air Blank		0.000	22:55
Control Test		0.080	22:55
Air Blank		0.000	22:56
Diagnostics Check		OK	22:56

Data Download Date/Time: 04/17/2019 10:58

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALMEIDA,CAMILO',J
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: ALMEIDA,CAMILO
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 04/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: GUTHRIE

DATE/TIME OF TEST: 04/01/2019 03:33:31

Diagnostics Check		OK	03:30	
Air Blank		0.000	03:31	
Control Test		0.080	03:31	
Air Blank		0.000	03:31	
Subject Sample	# 1		03:32	Subject Test Refused
Breath Volume		0.000	03:32	
Air Blank		0.000	03:32	
Control Test		0.079	03:32	
Air Blank		0.000	03:33	
Diagnostics Check		OK	03:33	

Data Download Date/Time: 04/24/2019 09:51

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULDERRIG,SEAN,F
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: MULDERRIG,SEAN
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 04/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 10:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:33

SUBJECT'S LAST NAME: WARLEN

DATE/TIME OF TEST: 04/04/2019 12:05:10

Diagnostics Check		OK	11:58
Air Blank		0.000	11:59
Control Test		0.079	11:59
Air Blank		0.000	12:00
Subject Sample	# 1	0.000	12:00
Breath Volume		3.644	12:00
Air Blank		0.000	12:01
Air Blank		0.000	12:03
Subject Sample	# 2	0.000	12:03
Breath Volume		3.390	12:03
Air Blank		0.000	12:04
Control Test		0.080	12:04
Air Blank		0.000	12:05
Diagnostics Check		OK	12:05

Data Download Date/Time: 04/24/2019 10:43

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULDERRIG,SEAN,F
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: MULDERRIG,SEAN
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 04/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 15:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:37

SUBJECT'S LAST NAME: BREWSTER

DATE/TIME OF TEST: 04/26/2019 16:05:38

Diagnostics Check		OK	15:59
Air Blank		0.000	15:59
Control Test		0.080	16:00
Air Blank		0.000	16:00
Subject Sample	# 1	0.095	16:01
Breath Volume		1.437	16:01
Air Blank		0.000	16:01
Air Blank		0.000	16:03
Subject Sample	# 2	0.088	16:04
Breath Volume		1.164	16:04
Air Blank		0.000	16:04
Control Test		0.079	16:05
Air Blank		0.000	16:05
Diagnostics Check		OK	16:05

Data Download Date/Time: 05/01/2019 10:37

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,DANIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LYONS,DANIEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 04/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: WARD

DATE/TIME OF TEST: 04/04/2019 00:50:17

Diagnostics Check		OK	00:44
Air Blank		0.000	00:44
Control Test		0.080	00:45
Air Blank		0.000	00:45
Subject Sample	# 1	0.000	00:46
Breath Volume		2.320	00:46
Air Blank		0.000	00:46
Air Blank		0.000	00:48
Subject Sample	# 2	0.000	00:48
Breath Volume		2.613	00:48
Air Blank		0.000	00:49
Control Test		0.080	00:49
Air Blank		0.000	00:50
Diagnostics Check		OK	00:50

Data Download Date/Time: 05/02/2019 05:10

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,DANIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LYONS,DANIEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 04/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: DEL VALLE TOLEDO

DATE/TIME OF TEST: 04/12/2019 20:34:51

Diagnostics Check		OK	20:20
Air Blank		0.000	20:21
Control Test		0.079	20:21
Air Blank		0.000	20:22
Subject Sample	# 1	0.194	20:23
Breath Volume		1.968	20:23
Air Blank		0.000	20:24
Air Blank		0.000	20:26
Subject Sample	# 2	0.171	20:29
Breath Volume		1.152	20:29
Air Blank		0.000	20:30
Air Blank		0.000	20:31
Subject Sample	# 3	0.184	20:33
Breath Volume		1.839	20:33
Air Blank		0.000	20:33
Control Test		0.075	20:34
Air Blank		0.000	20:34
Diagnostics Check		OK	20:34

Data Download Date/Time: 05/02/2019 05:10

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,DANIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LYONS,DANIEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 04/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: CONCEPCION RAMOS

DATE/TIME OF TEST: 04/13/2019 04:04:06

Diagnostics Check		OK	03:56
Air Blank		0.000	03:57
Control Test		0.078	03:57
Air Blank		0.000	03:58
Subject Sample	# 1	0.179	03:59
Breath Volume		1.640	03:59
Air Blank		0.000	04:00
Air Blank		0.000	04:01
Subject Sample	# 2	0.184	04:02
Breath Volume		1.484	04:02
Air Blank		0.000	04:03
Control Test		0.079	04:03
Air Blank		0.000	04:04
Diagnostics Check		OK	04:04

Data Download Date/Time: 05/02/2019 05:10

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,DANIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LYONS,DANIEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 04/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: SOTO

DATE/TIME OF TEST: 04/24/2019 00:41:01

Diagnostics Check		OK	00:32
Air Blank		0.000	00:32
Control Test		0.080	00:32
Air Blank		0.000	00:33
Subject Sample	# 1	0.195	00:36
Breath Volume		1.992	00:36
Air Blank		0.000	00:37
Air Blank		0.000	00:38
Subject Sample	# 2	0.189	00:39
Breath Volume		1.917	00:39
Air Blank		0.000	00:40
Control Test		0.076	00:40
Air Blank		0.000	00:40
Diagnostics Check		OK	00:41

Data Download Date/Time: 05/02/2019 05:10

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUFAHL,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PUFAHL,G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 03/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: OROZCO VASQUEZ

DATE/TIME OF TEST: 04/04/2019 03:01:58

Diagnostics Check		OK	02:55
Air Blank		0.000	02:56
Control Test		0.079	02:56
Air Blank		0.000	02:57
Subject Sample	# 1	0.207	02:57
Breath Volume		2.460	02:57
Air Blank		0.000	02:58
Air Blank		0.000	02:59
Subject Sample	# 2	0.225	03:00
Breath Volume		2.425	03:00
Air Blank		0.000	03:01
Control Test		0.078	03:01
Air Blank		0.000	03:01
Diagnostics Check		OK	03:01

Data Download Date/Time: 04/16/2019 08:57

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSTYAL,MATT,A
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KOSTYAL,MATT
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 03/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: SPEAKMAN

DATE/TIME OF TEST: 04/04/2019 23:47:21

Diagnostics Check		OK	23:41
Air Blank		0.000	23:41
Control Test		0.079	23:41
Air Blank		0.000	23:42
Subject Sample	# 1	0.169	23:42
Breath Volume		2.925	23:42
Air Blank		0.000	23:43
Air Blank		0.000	23:45
Subject Sample	# 2	0.167	23:45
Breath Volume		2.675	23:45
Air Blank		0.000	23:46
Control Test		0.078	23:46
Air Blank		0.000	23:47
Diagnostics Check		OK	23:47

Data Download Date/Time: 04/16/2019 08:57

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER,TYLER,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FOWLER,TYLER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 03/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: MILLERMON

DATE/TIME OF TEST: 04/05/2019 00:02:11

Diagnostics Check		OK	23:56
Air Blank		0.000	23:56
Control Test		0.078	23:56
Air Blank		0.000	23:57
Subject Sample	# 2	0.101	23:57
Breath Volume		4.125	23:57
Air Blank		0.000	23:58
Air Blank		0.000	00:00
Subject Sample	# 1	0.096	00:00
Breath Volume		4.289	00:00
Air Blank		0.000	00:01
Control Test		0.077	00:01
Air Blank		0.000	00:02
Diagnostics Check		OK	00:02

Data Download Date/Time: 04/16/2019 08:57

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUFAHL,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PUFAHL,G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 03/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 04/06/2019 03:08:33

Diagnostics Check		OK	02:59
Air Blank		0.000	03:00
Control Test		0.079	03:00
Air Blank		0.000	03:01
Subject Sample	# 1	0.168	03:03
Breath Volume		1.335	03:03
Air Blank		0.000	03:03
Air Blank		0.000	03:05
Subject Sample	# 2	0.158	03:06
Breath Volume		1.171	03:06
Air Blank		0.000	03:07
Control Test		0.078	03:07
Air Blank		0.000	03:08
Diagnostics Check		OK	03:08

Data Download Date/Time: 04/16/2019 08:57

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSTYAL,MATT,A
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KOSTYAL,MATT
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 03/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:58

SUBJECT'S LAST NAME: BRANDT

DATE/TIME OF TEST: 04/10/2019 19:26:13

Diagnostics Check		OK	19:19
Air Blank		0.000	19:20
Control Test		0.078	19:20
Air Blank		0.000	19:20
Subject Sample	# 1	0.123	19:21
Breath Volume		2.871	19:21
Air Blank		0.000	19:22
Air Blank		0.000	19:23
Subject Sample	# 2	0.122	19:24
Breath Volume		3.578	19:24
Air Blank		0.000	19:25
Control Test		0.077	19:25
Air Blank		0.000	19:26
Diagnostics Check		OK	19:26

Data Download Date/Time: 04/16/2019 08:57

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUEHLER',DEP,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BUEHLER,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 04/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: CRUZ CAMPILLO

DATE/TIME OF TEST: 04/20/2019 02:57:34

Diagnostics Check		OK	02:50
Air Blank		0.000	02:50
Control Test		0.079	02:50
Air Blank		0.000	02:51
Subject Sample	# 1	0.213	02:52
Breath Volume		1.609	02:52
Air Blank		0.000	02:53
Air Blank		0.000	02:55
Subject Sample	# 2	0.221	02:55
Breath Volume		1.937	02:55
Air Blank		0.000	02:56
Control Test		0.078	02:56
Air Blank		0.000	02:57
Diagnostics Check		OK	02:57

Data Download Date/Time: 05/28/2019 10:02

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUFAHL,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PUFAHL,G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 04/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: TRUELOVE

DATE/TIME OF TEST: 04/24/2019 22:27:13

Diagnostics Check		OK	22:15	
Air Blank		0.000	22:15	
Control Test		0.079	22:16	
Air Blank		0.000	22:16	
Subject Sample	# 1	0.278	22:19	Volume Not Met
Breath Volume		0.902	22:19	
Air Blank		0.000	22:20	
Air Blank		0.000	22:22	
Subject Sample	# 2	0.281	22:25	Volume Not Met
Breath Volume		0.816	22:25	
Air Blank		0.000	22:26	
Control Test		0.075	22:26	
Air Blank		0.000	22:27	
Diagnostics Check		OK	22:27	

Data Download Date/Time: 05/28/2019 10:02

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUFAHL,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PUFAHL,G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 04/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:42

SUBJECT'S LAST NAME: STAMBAUGH

DATE/TIME OF TEST: 04/25/2019 20:09:36

Diagnostics Check		OK	20:02
Air Blank		0.000	20:03
Control Test		0.077	20:03
Air Blank		0.000	20:04
Subject Sample	# 1	0.173	20:04
Breath Volume		2.363	20:04
Air Blank		0.000	20:05
Air Blank		0.000	20:07
Subject Sample	# 2	0.174	20:07
Breath Volume		3.000	20:07
Air Blank		0.000	20:08
Control Test		0.078	20:08
Air Blank		0.000	20:09
Diagnostics Check		OK	20:09

Data Download Date/Time: 05/28/2019 10:02

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUFAHL,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PUFAHL,G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 04/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: HUTCHINSON

DATE/TIME OF TEST: 04/27/2019 00:05:29

Diagnostics Check		OK	23:54	
Air Blank		0.000	23:54	
Control Test		0.079	23:54	
Air Blank		0.000	23:55	
Subject Sample	# 2		23:58	No Sample Provided
Breath Volume		0.000	23:58	
Air Blank		0.000	23:58	
Air Blank		0.000	00:00	
Subject Sample	# 1		00:04	No Sample Provided
Breath Volume		0.000	00:04	
Air Blank		0.000	00:04	
Control Test		0.078	00:04	
Air Blank		0.000	00:05	
Diagnostics Check		OK	00:05	

Data Download Date/Time: 05/28/2019 10:02

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUFAHL,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PUFAHL,G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 04/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: SMALL

DATE/TIME OF TEST: 04/27/2019 23:13:59

Diagnostics Check		OK	23:05
Air Blank		0.000	23:05
Control Test		0.079	23:05
Air Blank		0.000	23:06
Subject Sample	# 1	0.191	23:08
Breath Volume		1.289	23:08
Air Blank		0.000	23:09
Air Blank		0.000	23:10
Subject Sample	# 2	0.184	23:12
Breath Volume		1.648	23:12
Air Blank		0.000	23:13
Control Test		0.077	23:13
Air Blank		0.000	23:13
Diagnostics Check		OK	23:13

Data Download Date/Time: 05/28/2019 10:02

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUFAHL,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PUFAHL,G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 04/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: FRUDAKIS

DATE/TIME OF TEST: 04/30/2019 22:28:41

Diagnostics Check		OK	22:17	
Air Blank		0.000	22:17	
Control Test		0.078	22:18	
Air Blank		0.000	22:18	
Subject Sample	# 1		22:21	No Sample Provided
Breath Volume		0.000	22:21	
Air Blank		0.000	22:22	
Air Blank		0.000	22:24	
Subject Sample	# 2		22:27	No Sample Provided
Breath Volume		0.000	22:27	
Air Blank		0.000	22:27	
Control Test		0.079	22:28	
Air Blank		0.000	22:28	
Diagnostics Check		OK	22:28	

Data Download Date/Time: 05/28/2019 10:02

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FREY,JACOB
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:29

SUBJECT'S LAST NAME: WEISZ

DATE/TIME OF TEST: 04/14/2019 18:36:52

Diagnostics Check		OK	18:30
Air Blank		0.000	18:30
Control Test		0.079	18:30
Air Blank		0.000	18:31
Subject Sample	# 1	0.122	18:32
Breath Volume		2.351	18:32
Air Blank		0.000	18:32
Air Blank		0.000	18:34
Subject Sample	# 2	0.124	18:35
Breath Volume		3.304	18:35
Air Blank		0.000	18:35
Control Test		0.077	18:36
Air Blank		0.000	18:36
Diagnostics Check		OK	18:36

Data Download Date/Time: 04/19/2019 13:43

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ZEITZ,PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: BUI
ARREST TIME: 17:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: VENZARA

DATE/TIME OF TEST: 04/14/2019 19:24:05

Diagnostics Check		OK	19:16
Air Blank		0.000	19:17
Control Test		0.078	19:17
Air Blank		0.000	19:17
Subject Sample	# 1	0.155	19:18
Breath Volume		4.425	19:18
Air Blank		0.000	19:19
Air Blank		0.000	19:21
Subject Sample	# 2	0.154	19:22
Breath Volume		3.925	19:22
Air Blank		0.000	19:23
Control Test		0.078	19:23
Air Blank		0.000	19:24
Diagnostics Check		OK	19:24

Data Download Date/Time: 04/19/2019 13:43

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DEFEO,PATRICK
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: BUI
ARREST TIME: 18:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:58

SUBJECT'S LAST NAME: QUINTANA

DATE/TIME OF TEST: 04/14/2019 20:17:56

Diagnostics Check		OK	20:11
Air Blank		0.000	20:11
Control Test		0.078	20:12
Air Blank		0.000	20:12
Subject Sample	# 1	0.157	20:13
Breath Volume		2.367	20:13
Air Blank		0.000	20:13
Air Blank		0.000	20:15
Subject Sample	# 2	0.159	20:16
Breath Volume		2.878	20:16
Air Blank		0.000	20:17
Control Test		0.078	20:17
Air Blank		0.000	20:17
Diagnostics Check		OK	20:17

Data Download Date/Time: 04/19/2019 13:43

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006046

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARTNER,ARTHUR,F
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GARTNER,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1048227
DRY GAS EXPIRY DATE: 11/15/2021
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: ROLLOCKS

DATE/TIME OF TEST: 04/13/2019 00:40:14

Diagnostics Check		OK	00:32
Air Blank		0.000	00:33
Control Test		0.078	00:33
Air Blank		0.000	00:34
Subject Sample	# 1	0.144	00:35
Breath Volume		1.617	00:35
Air Blank		0.000	00:35
Air Blank		0.000	00:37
Subject Sample	# 2	0.140	00:38
Breath Volume		1.414	00:38
Air Blank		0.000	00:39
Control Test		0.078	00:39
Air Blank		0.000	00:40
Diagnostics Check		OK	00:40

Data Download Date/Time: 05/02/2019 17:32

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006046

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,TABARIE,G
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SULLIVAN,TABARIE
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1048227
DRY GAS EXPIRY DATE: 11/15/2021
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: CZEPIEL

DATE/TIME OF TEST: 04/24/2019 03:43:19

Diagnostics Check		OK	03:36
Air Blank		0.000	03:36
Control Test		0.079	03:36
Air Blank		0.000	03:37
Subject Sample	# 1	0.086	03:37
Breath Volume		1.316	03:37
Air Blank		0.000	03:38
Air Blank		0.000	03:40
Subject Sample	# 2	0.093	03:41
Breath Volume		1.972	03:41
Air Blank		0.000	03:42
Control Test		0.081	03:42
Air Blank		0.000	03:43
Diagnostics Check		OK	03:43

Data Download Date/Time: 05/02/2019 17:32

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006085

INSTRUMENT REGISTERED TO: US COAST GUARD MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARNER,TYLER,E
BREATH TEST OPERATOR AGENCY: US COAST GUARD MIAMI
ARRESTING OFFICER: CARTAYA,GUILLERMO
ARRESTING OFFICER AGENCY: FFWCC (SOUTH A)
LAST AGENCY INSPECTION: 03/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: BUI
ARREST TIME: 21:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:03

SUBJECT'S LAST NAME: LUIS

DATE/TIME OF TEST: 04/07/2019 22:14:18

Diagnostics Check		OK	22:05	
Air Blank		0.000	22:06	
Control Test		0.080	22:06	
Air Blank		0.000	22:07	
Subject Sample	# 1		22:10	No Sample Provided
Breath Volume		0.000	22:10	
Air Blank		0.000	22:10	
Air Blank		0.000	22:12	
Subject Sample	# 2		22:12	Subject Test Refused
Breath Volume		0.000	22:12	
Air Blank		0.000	22:13	
Control Test		0.080	22:13	
Air Blank		0.000	22:14	
Diagnostics Check		OK	22:14	

Data Download Date/Time: 04/26/2019 11:43

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.