Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PILKANIS,MARY,C
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: WAGNER,RYAN
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 04/01/2019 19:18:46

Diagnostics Check		OK	19:15	
Air Blank		0.000	19:16	
Control Test		0.080	19:16	
Air Blank		0.000	19:17	
Subject Sample	# 1		19:17	Subject Test Refused
Breath Volume		0.000	19:17	
Air Blank		0.000	19:17	
Control Test		0.081	19:18	
Air Blank		0.000	19:18	
Diagnostics Check		OK	19:18	

Data Download Date/Time: 04/22/2019 13:30

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLER,DAVID,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: KELLER,DAVID
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: FREEMAN

DATE/TIME OF TEST: 04/04/2019 00:12:08

Diagnostics Check		OK	00:05
Air Blank		0.000	00:06
Control Test		0.081	00:06
Air Blank		0.000	00:07
Subject Sample	# 1	0.126	00:07
Breath Volume		2.125	00:07
Air Blank		0.000	00:08
Air Blank		0.000	00:10
Subject Sample	# 2	0.132	00:10
Breath Volume		2.640	00:10
Air Blank		0.000	00:11
Control Test		0.080	00:11
Air Blank		0.000	00:12
Diagnostics Check		OK	00:12

Data Download Date/Time: 04/22/2019 13:30

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITAKER,HUNTER,H
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: WHITAKER,HUNTER
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: SHAW

DATE/TIME OF TEST: 04/05/2019 03:34:28

Diagnostics Check		OK	03:27
Air Blank		0.000	03:28
Control Test		0.081	03:28
Air Blank		0.000	03:29
Subject Sample	# 1	0.094	03:29
Breath Volume		3.898	03:29
Air Blank		0.000	03:30
Air Blank		0.000	03:32
Subject Sample	# 2	0.084	03:32
Breath Volume		2.222	03:32
Air Blank		0.000	03:33
Control Test		0.081	03:33
Air Blank		0.000	03:34
Diagnostics Check		OK	03:34

Data Download Date/Time: 04/22/2019 13:30

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FINUFF,DANIEL,W
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: FINUFF,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: LOVELADY

DATE/TIME OF TEST: 04/06/2019 02:15:43

Diagnostics Check		OK	02:09
Air Blank		0.000	02:10
Control Test		0.081	02:10
Air Blank		0.000	02:10
Subject Sample	# 1	0.000	02:11
Breath Volume		1.304	02:11
Air Blank		0.000	02:11
Air Blank		0.000	02:13
Subject Sample	# 2	0.000	02:14
Breath Volume		2.386	02:14
Air Blank		0.000	02:14
Control Test		0.081	02:15
Air Blank		0.000	02:15
Diagnostics Check		OK	02:15

Data Download Date/Time: 04/22/2019 13:30

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWINDLE,CHARLES,M
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: SWINDLE,CHARLES
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: SENATUS

DATE/TIME OF TEST: 04/08/2019 22:11:34

Diagnostics Check		OK	22:05
Air Blank		0.000	22:06
Control Test		0.081	22:06
Air Blank		0.000	22:06
Subject Sample	# 1	0.000	22:07
Breath Volume		1.851	22:07
Air Blank		0.000	22:07
Air Blank		0.000	22:09
Subject Sample	# 2	0.000	22:10
Breath Volume		2.007	22:10
Air Blank		0.000	22:10
Control Test		0.081	22:10
Air Blank		0.000	22:11
Diagnostics Check		OK	22:11

Data Download Date/Time: 04/22/2019 13:30

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLER,DAVID,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: MIMS,C
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: BUI
ARREST TIME: 20:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: BARWICK

DATE/TIME OF TEST: 04/10/2019 22:59:51

Diagnostics Check		OK	22:52
Air Blank		0.000	22:52
Control Test		0.081	22:53
Air Blank		0.000	22:53
Subject Sample	# 1	0.079	22:54
Breath Volume		1.300	22:54
Air Blank		0.000	22:54
Air Blank		0.000	22:56
Subject Sample	# 2	0.079	22:58
Breath Volume		1.636	22:58
Air Blank		0.000	22:58
Control Test		0.082	22:59
Air Blank		0.000	22:59
Diagnostics Check		OK	22:59

Data Download Date/Time: 04/22/2019 13:30

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARVEY,RHETT,A
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: HARVEY,RHETT
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: MORROW

DATE/TIME OF TEST: 04/12/2019 01:55:51

Diagnostics Check		OK	01:49
Air Blank		0.000	01:49
Control Test		0.081	01:50
Air Blank		0.000	01:50
Subject Sample	# 1	0.156	01:51
Breath Volume		3.515	01:51
Air Blank		0.000	01:51
Air Blank		0.000	01:53
Subject Sample	# 2	0.172	01:54
Breath Volume		3.496	01:54
Air Blank		0.000	01:54
Control Test		0.081	01:55
Air Blank		0.000	01:55
Diagnostics Check		OK	01:55

Data Download Date/Time: 04/22/2019 13:30

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,PARISE,L
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: ADAMS,PARISE
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: GROLL

DATE/TIME OF TEST: 04/13/2019 01:22:10

Diagnostics Check		OK	01:12	
Air Blank		0.000	01:12	
Control Test		0.080	01:13	
Air Blank		0.000	01:13	
Subject Sample	# 1	0.197	01:14	
Breath Volume		1.175	01:14	
Air Blank		0.000	01:15	
Air Blank		0.000	01:17	
Subject Sample	# 2	0.204	01:20	Volume Not Met
Breath Volume		0.050	01:20	
Air Blank			01:21	Purge Fail
Air Blank			01:22	Purge Fail

Data Download Date/Time: 04/22/2019 13:30

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,PARISE,L
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: ADAMS,PARISE
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: GROLL

DATE/TIME OF TEST: 04/13/2019 01:25:43

Diagnostics Check	OK	01:24	
Air Blank	0.000	01:24	
Control Test	0.050	01:25	Control Outside Tolerance
Air Blank	0.000	01:25	

Data Download Date/Time: 04/22/2019 13:30

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,PARISE,L
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: ADAMS,PARISE
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: GROLL

DATE/TIME OF TEST: 04/13/2019 01:28:43

Diagnostics Check	OK	01:27	
Air Blank	0.000	01:27	
Control Test	0.061	01:28	Control Outside Tolerance
Air Blank	0.000	01:28	

Data Download Date/Time: 04/22/2019 13:30

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILCOX,DAVID,B
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: PHILCOX,DAVID
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: SWANGO

DATE/TIME OF TEST: 04/18/2019 03:34:08

Diagnostics Check		OK	03:26
Air Blank		0.000	03:27
Control Test		0.080	03:27
Air Blank		0.000	03:27
Subject Sample	# 1	0.027	03:28
Breath Volume		1.101	03:28
Air Blank		0.000	03:29
Air Blank		0.000	03:31
Subject Sample	# 2	0.025	03:32
Breath Volume		1.894	03:32
Air Blank		0.000	03:33
Control Test		0.080	03:33
Air Blank		0.000	03:34
Diagnostics Check		OK	03:34

Data Download Date/Time: 04/22/2019 13:30

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINSON,DARIN,L
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: ROBINSON,DARIN
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: KENNEDY

DATE/TIME OF TEST: 04/19/2019 23:47:12

Diagnostics Check	OK	23:46	
Air Blank	0.000	23:46	
Control Test	0.000	23:46	RFI Detect
Air Blank	0.000	23:47	

Data Download Date/Time: 04/22/2019 13:30

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINSON,DARIN,L
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: ROBINSON,DARIN
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: KENNEDY

DATE/TIME OF TEST: 04/19/2019 23:59:01

Diagnostics Check		OK	23:52
Air Blank		0.000	23:52
Control Test		0.081	23:53
Air Blank		0.000	23:53
Subject Sample	# 1	0.211	23:54
Breath Volume		1.207	23:54
Air Blank		0.000	23:54
Air Blank		0.000	23:56
Subject Sample	# 2	0.213	23:57
Breath Volume		1.539	23:57
Air Blank		0.000	23:58
Control Test		0.081	23:58
Air Blank		0.000	23:58
Diagnostics Check		OK	23:59

Data Download Date/Time: 04/22/2019 13:30

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUNTER,ROBERT,A
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: WESTMAN,PATRICK
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 04/20/2019 01:36:25

Diagnostics Check		OK	01:30
Air Blank		0.000	01:30
Control Test		0.081	01:30
Air Blank		0.000	01:31
Subject Sample	# 1	0.187	01:32
Breath Volume		2.011	01:32
Air Blank		0.000	01:32
Air Blank		0.000	01:34
Subject Sample	# 2	0.173	01:34
Breath Volume		1.457	01:34
Air Blank		0.000	01:35
Control Test		0.081	01:35
Air Blank		0.000	01:36
Diagnostics Check		OK	01:36

Data Download Date/Time: 04/22/2019 13:30

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAXTON,D,L
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: BRAXTON,D
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: DUDLEY

DATE/TIME OF TEST: 04/20/2019 03:46:36

Diagnostics Check		OK	03:39
Air Blank		0.000	03:39
Control Test		0.081	03:39
Air Blank		0.000	03:40
Subject Sample	# 1	0.148	03:41
Breath Volume		2.332	03:41
Air Blank		0.000	03:42
Air Blank		0.000	03:44
Subject Sample	# 2	0.136	03:45
Breath Volume		1.445	03:45
Air Blank		0.000	03:45
Control Test		0.079	03:46
Air Blank		0.000	03:46
Diagnostics Check		OK	03:46

Data Download Date/Time: 04/22/2019 13:30

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCARTHY,TED,G
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: PEART,W
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 11:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:47

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 04/20/2019 12:12:28

Diagnostics Check		OK	12:09	
Air Blank		0.000	12:10	
Control Test		0.081	12:10	
Air Blank		0.000	12:10	
Subject Sample	# 1		12:11	Subject Test Refused
Breath Volume		0.000	12:11	
Air Blank		0.000	12:11	
Control Test		0.081	12:11	
Air Blank		0.000	12:12	
Diagnostics Check		OK	12:12	

Data Download Date/Time: 04/22/2019 13:30

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEELE,R,L
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: STEELE,R
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 04/21/2019 02:21:52

Diagnostics Check		OK	02:15
Air Blank		0.000	02:15
Control Test		0.081	02:15
Air Blank		0.000	02:16
Subject Sample	# 1	0.213	02:17
Breath Volume		2.968	02:17
Air Blank		0.000	02:17
Air Blank		0.000	02:19
Subject Sample	# 2	0.228	02:20
Breath Volume		3.652	02:20
Air Blank		0.000	02:20
Control Test		0.080	02:21
Air Blank		0.000	02:21
Diagnostics Check		OK	02:21

Data Download Date/Time: 04/22/2019 13:30

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLER,DAVID,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: KELLER,DAVID
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 04/21/2019 04:48:28

Diagnostics Check		OK	04:42
Air Blank		0.000	04:42
Control Test		0.081	04:43
Air Blank		0.000	04:43
Subject Sample	# 1	0.126	04:44
Breath Volume		2.585	04:44
Air Blank		0.000	04:44
Air Blank		0.000	04:46
Subject Sample	# 2	0.123	04:47
Breath Volume		2.253	04:47
Air Blank		0.000	04:47
Control Test		0.081	04:47
Air Blank		0.000	04:48
Diagnostics Check		OK	04:48

Data Download Date/Time: 04/22/2019 13:30

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLLINS,C,P
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: DEANTHONY,S
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 04/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 06:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:50

SUBJECT'S LAST NAME: MANNINGS

DATE/TIME OF TEST: 04/27/2019 07:23:25

Diagnostics Check		OK	07:15
Air Blank		0.000	07:15
Control Test		0.081	07:16
Air Blank		0.000	07:16
Subject Sample	# 1	0.210	07:18
Breath Volume		1.210	07:18
Air Blank		0.000	07:19
Air Blank		0.000	07:20
Subject Sample	# 2	0.201	07:21
Breath Volume		1.562	07:21
Air Blank		0.000	07:22
Control Test		0.080	07:22
Air Blank		0.000	07:23
Diagnostics Check		OK	07:23

Data Download Date/Time: 05/21/2019 10:57

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,GARY,W
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: HOWARD,BLAKE
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 04/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 12:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:17

SUBJECT'S LAST NAME: CROSBY

DATE/TIME OF TEST: 04/27/2019 13:49:58

Diagnostics Check		OK	13:43
Air Blank		0.000	13:44
Control Test		0.081	13:44
Air Blank		0.000	13:45
Subject Sample	# 1	0.224	13:45
Breath Volume		1.609	13:45
Air Blank		0.000	13:46
Air Blank		0.000	13:48
Subject Sample	# 2	0.216	13:48
Breath Volume		1.703	13:48
Air Blank		0.000	13:49
Control Test		0.080	13:49
Air Blank		0.000	13:49
Diagnostics Check		OK	13:49

Data Download Date/Time: 05/21/2019 10:57

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSCITTO,C,E
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: RUSCITTO,C
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:59

SUBJECT'S LAST NAME: POLZER

DATE/TIME OF TEST: 04/07/2019 22:38:33

Diagnostics Check		OK	22:29
Air Blank		0.000	22:29
Control Test		0.081	22:30
Air Blank		0.000	22:30
Subject Sample	# 1	0.233	22:31
Breath Volume		2.097	22:31
Air Blank		0.000	22:31
Air Blank		0.000	22:33
Subject Sample	# 2	0.210	22:34
Breath Volume		1.132	22:34
Air Blank		0.000	22:34
Air Blank		0.000	22:36
Subject Sample	# 3	0.229	22:37
Breath Volume		1.722	22:37
Air Blank		0.000	22:37
Control Test		0.079	22:37
Air Blank		0.000	22:38
Diagnostics Check		OK	22:38

Data Download Date/Time: 04/22/2019 13:37

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAN,DONALD,E
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: DEAN,DONALD
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:26

SUBJECT'S LAST NAME: SAFFORD

DATE/TIME OF TEST: 04/09/2019 20:58:25

Diagnostics Check		OK	20:52	
Air Blank		0.000	20:52	
Control Test		0.081	20:53	
Air Blank		0.000	20:53	
Subject Sample	# 1	0.000	20:56	Volume Not Met
Breath Volume		0.468	20:56	
Air Blank			20:57	RFI Detect
Air Blank		0.000	20:57	
Control Test		0.081	20:57	
Air Blank		0.000	20:58	
Diagnostics Check		OK	20:58	

Data Download Date/Time: 04/22/2019 13:37

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUNTER,ROBERT,A
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: WESTMAN,PATRICK
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 04/20/2019 01:22:46

Diagnostics Check		OK	01:13
Air Blank		0.000	01:13
Control Test		0.079	01:14
Air Blank		0.000	01:14
Subject Sample	# 1	0.193	01:15
Breath Volume		2.339	01:15
Air Blank		0.000	01:15
Air Blank		0.000	01:17
Subject Sample	# 2	0.166	01:18
Breath Volume		2.058	01:18
Air Blank		0.000	01:18
Air Blank		0.000	01:20
Subject Sample	# 3	0.186	01:21
Breath Volume		2.265	01:21
Air Blank		0.000	01:21
Control Test		0.080	01:22
Air Blank		0.000	01:22
Diagnostics Check		OK	01:22

Data Download Date/Time: 04/22/2019 13:37

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINSON,DARIN,L
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: ROBINSON,DARIN
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 04/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 04/28/2019 23:33:37

Diagnostics Check		OK	23:25
Air Blank		0.000	23:25
Control Test		0.081	23:26
Air Blank		0.000	23:26
Subject Sample	# 1	0.178	23:29
Breath Volume		2.191	23:29
Air Blank		0.000	23:29
Air Blank		0.000	23:31
Subject Sample	# 2	0.178	23:32
Breath Volume		2.031	23:32
Air Blank		0.000	23:32
Control Test		0.079	23:33
Air Blank		0.000	23:33
Diagnostics Check		OK	23:33

Data Download Date/Time: 05/21/2019 11:34

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLLINS,CAMERON,P
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: COLLINS,CAMERON
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 04/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: RITTMAN 3

DATE/TIME OF TEST: 04/30/2019 02:02:59

Diagnostics Check		OK	01:56
Air Blank		0.000	01:56
Control Test		0.081	01:57
Air Blank		0.000	01:57
Subject Sample	# 1	0.117	01:58
Breath Volume		1.574	01:58
Air Blank		0.000	01:58
Air Blank		0.000	02:00
Subject Sample	# 2	0.121	02:01
Breath Volume		1.308	02:01
Air Blank		0.000	02:02
Control Test		0.081	02:02
Air Blank		0.000	02:02
Diagnostics Check		OK	02:02

Data Download Date/Time: 05/21/2019 11:34

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYMONETTE,ASHLEY,N
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: LITZKOW,CARLTON
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 03/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: SYMONETTE

DATE/TIME OF TEST: 04/06/2019 04:17:55

Diagnostics Check		OK	04:02	
Air Blank		0.000	04:02	
Control Test		0.080	04:02	
Air Blank		0.000	04:03	
Subject Sample	# 1	0.220	04:06	Volume Not Met
Breath Volume		0.890	04:06	
Air Blank		0.000	04:07	
Air Blank		0.000	04:08	
Subject Sample	# 2	0.257	04:10	
Breath Volume		1.910	04:10	
Air Blank		0.000	04:11	
Air Blank		0.000	04:13	
Subject Sample	# 3		04:16	No Sample Provided
Breath Volume		0.000	04:16	
Air Blank		0.000	04:17	
Control Test		0.076	04:17	
Air Blank		0.000	04:17	
Diagnostics Check		OK	04:17	

Data Download Date/Time: 04/25/2019 10:03

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FUNDENBURG,TRACY,C
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STANFIELD,THOMAS
ARRESTING OFFICER AGENCY: Alachua PD
LAST AGENCY INSPECTION: 03/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 07:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:21

SUBJECT'S LAST NAME: BURNETT

DATE/TIME OF TEST: 04/07/2019 07:52:38

Diagnostics Check		OK	07:48	
Air Blank		0.000	07:48	
Control Test		0.081	07:48	
Air Blank		0.000	07:49	
Subject Sample	# 1		07:52	Sequence Aborted
Breath Volume		0.000	07:52	
Air Blank		0.000	07:52	

Data Download Date/Time: 04/25/2019 10:03

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LITZKOW,CLAYTON,P
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: LITZKOW,CLAYTON
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 03/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: SPOONE

DATE/TIME OF TEST: 04/11/2019 03:50:38

Diagnostics Check		OK	03:44
Air Blank		0.000	03:44
Control Test		0.079	03:45
Air Blank		0.000	03:45
Subject Sample	# 1	0.215	03:46
Breath Volume		1.835	03:46
Air Blank		0.000	03:46
Air Blank		0.000	03:48
Subject Sample	# 2	0.209	03:49
Breath Volume		1.773	03:49
Air Blank		0.000	03:49
Control Test		0.076	03:50
Air Blank		0.000	03:50
Diagnostics Check		OK	03:50

Data Download Date/Time: 04/25/2019 10:03

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURINGTON,PATRICK,A
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: PURINGTON,PATRICK
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 03/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: SAEMANN

DATE/TIME OF TEST: 04/13/2019 19:32:50

Diagnostics Check		OK	19:26
Air Blank		0.000	19:26
Control Test		0.078	19:26
Air Blank		0.000	19:27
Subject Sample	# 1	0.164	19:28
Breath Volume		2.378	19:28
Air Blank		0.000	19:29
Air Blank		0.000	19:30
Subject Sample	# 2	0.170	19:31
Breath Volume		2.984	19:31
Air Blank		0.000	19:31
Control Test		0.077	19:32
Air Blank		0.000	19:32
Diagnostics Check		OK	19:32

Data Download Date/Time: 04/25/2019 10:03

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUTSCHER,REBECCA,L
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: RIVERA,LEWIS
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 03/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: BISER

DATE/TIME OF TEST: 04/13/2019 23:22:37

Diagnostics Check		OK	23:14
Air Blank		0.000	23:14
Control Test		0.080	23:15
Air Blank		0.000	23:15
Subject Sample	# 1	0.206	23:17
Breath Volume		1.656	23:17
Air Blank		0.000	23:18
Air Blank		0.000	23:20
Subject Sample	# 2	0.189	23:20
Breath Volume		1.230	23:20
Air Blank		0.000	23:21
Control Test		0.076	23:22
Air Blank		0.000	23:22
Diagnostics Check		OK	23:22

Data Download Date/Time: 04/25/2019 10:03

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURINGTON,PATRICK,A
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: PURINGTON,PATRICK
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 03/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: FOSTER

DATE/TIME OF TEST: 04/14/2019 01:07:50

Diagnostics Check		OK	00:56	
Air Blank		0.000	00:56	
Control Test		0.080	00:57	
Air Blank		0.000	00:57	
Subject Sample	# 1		01:00	No Sample Provided
Breath Volume		0.000	01:00	
Air Blank		0.000	01:01	
Air Blank		0.000	01:03	
Subject Sample	# 2		01:06	No Sample Provided
Breath Volume		0.000	01:06	
Air Blank		0.000	01:06	
Control Test		0.081	01:07	
Air Blank		0.000	01:07	
Diagnostics Check		OK	01:07	

Data Download Date/Time: 04/25/2019 10:03

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURINGTON,PATRICK,A
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: PURINGTON,PATRICK
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 03/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: FAIRLEY

DATE/TIME OF TEST: 04/15/2019 03:54:22

Diagnostics Check		OK	03:42	
Air Blank		0.000	03:43	
Control Test		0.079	03:43	
Air Blank		0.000	03:44	
Subject Sample	# 1		03:47	No Sample Provided
Breath Volume		0.000	03:47	
Air Blank		0.000	03:47	
Air Blank		0.000	03:49	
Subject Sample	# 2		03:53	No Sample Provided
Breath Volume		0.000	03:53	
Air Blank		0.000	03:53	
Control Test		0.081	03:53	
Air Blank		0.000	03:54	
Diagnostics Check		OK	03:54	

Data Download Date/Time: 04/25/2019 10:03

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRICKER,JEFFREY,S
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: FRICKER,JEFFREY
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 03/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 05:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:30

SUBJECT'S LAST NAME: STEPHENS

DATE/TIME OF TEST: 04/19/2019 07:13:05

Diagnostics Check		OK	07:04	
Air Blank		0.000	07:04	
Control Test		0.079	07:05	
Air Blank		0.000	07:05	
Subject Sample	# 1		07:08	No Sample Provided
Breath Volume		0.000	07:08	
Air Blank		0.000	07:09	
Air Blank		0.000	07:11	
Subject Sample	# 2		07:11	Subject Test Refused
Breath Volume		0.000	07:11	
Air Blank		0.000	07:12	
Control Test		0.080	07:12	
Air Blank		0.000	07:12	
Diagnostics Check		OK	07:13	

Data Download Date/Time: 04/25/2019 10:03

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LITZKOW,CLAYTON,P
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: LITZKOW,CLAYTON
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 03/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: MACMANUS

DATE/TIME OF TEST: 04/22/2019 00:40:04

Diagnostics Check		OK	00:33
Air Blank		0.000	00:34
Control Test		0.080	00:34
Air Blank		0.000	00:34
Subject Sample	# 1	0.165	00:35
Breath Volume		1.976	00:35
Air Blank		0.000	00:36
Air Blank		0.000	00:38
Subject Sample	# 2	0.162	00:38
Breath Volume		1.867	00:38
Air Blank		0.000	00:39
Control Test		0.077	00:39
Air Blank		0.000	00:39
Diagnostics Check		OK	00:40

Data Download Date/Time: 04/25/2019 10:03

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVY,JAYSON,I
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: COONCE,JAMES
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 04/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 13:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:38

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 04/25/2019 14:06:16

Diagnostics Check		OK	13:59
Air Blank		0.000	14:00
Control Test		0.079	14:00
Air Blank		0.000	14:01
Subject Sample	# 1	0.313	14:01
Breath Volume		2.191	14:01
Air Blank		0.000	14:02
Air Blank		0.000	14:04
Subject Sample	# 2	0.316	14:04
Breath Volume		2.371	14:04
Air Blank		0.000	14:05
Control Test		0.078	14:05
Air Blank		0.000	14:06
Diagnostics Check		OK	14:06

Data Download Date/Time: 05/22/2019 10:08

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURINGTON,PATRICK,A
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: PURINGTON,PATRICK
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 04/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 04/28/2019 03:42:33

Diagnostics Check		OK	03:36
Air Blank		0.000	03:36
Control Test		0.080	03:37
Air Blank		0.000	03:37
Subject Sample	# 1	0.159	03:38
Breath Volume		1.796	03:38
Air Blank		0.000	03:38
Air Blank		0.000	03:40
Subject Sample	# 2	0.160	03:41
Breath Volume		1.203	03:41
Air Blank		0.000	03:41
Control Test		0.079	03:41
Air Blank		0.000	03:42
Diagnostics Check		OK	03:42

Data Download Date/Time: 05/22/2019 10:08

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESNEUL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LEWIS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 17:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 04/02/2019 19:11:53

Diagnostics Check	OK	19:10	
Air Blank	0.000	19:10	
Control Test	0.073	19:11	Control Outside Tolerance
Air Blank	0.000	19:11	

Data Download Date/Time: 04/29/2019 15:18

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESNEULM,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LEWIS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 17:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 04/02/2019 19:20:09

Diagnostics Check		OK	19:13	
Air Blank		0.000	19:13	
Control Test		0.076	19:14	
Air Blank		0.000	19:14	
Subject Sample	# 1	0.000	19:15	
Breath Volume		3.695	19:15	
Air Blank		0.000	19:16	
Air Blank		0.000	19:18	
Subject Sample	# 2	0.000	19:18	
Breath Volume		3.335	19:18	
Air Blank		0.000	19:19	
Control Test		0.073	19:19	Control Outside Tolerance
Air Blank		0.000	19:20	

Data Download Date/Time: 04/29/2019 15:18

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,KOURTNEY,B
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: OVERBAY,WILLIAM
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 05:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:35

SUBJECT'S LAST NAME: COSSON

DATE/TIME OF TEST: 04/11/2019 06:01:43

Diagnostics Check		OK	05:55
Air Blank		0.000	05:56
Control Test		0.081	05:56
Air Blank		0.000	05:56
Subject Sample	# 1	0.000	05:57
Breath Volume		1.718	05:57
Air Blank		0.000	05:57
Air Blank		0.000	05:59
Subject Sample	# 2	0.000	06:00
Breath Volume		1.464	06:00
Air Blank		0.000	06:00
Control Test		0.078	06:01
Air Blank		0.000	06:01
Diagnostics Check		OK	06:01

Data Download Date/Time: 04/29/2019 15:18

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURGESS,MASON,W
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BURGESS,MASON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: RICHESON

DATE/TIME OF TEST: 04/13/2019 02:01:14

Diagnostics Check		OK	01:54
Air Blank		0.000	01:54
Control Test		0.078	01:54
Air Blank		0.000	01:55
Subject Sample	# 1	0.219	01:56
Breath Volume		1.886	01:56
Air Blank		0.000	01:57
Air Blank		0.000	01:58
Subject Sample	# 2	0.208	01:59
Breath Volume		1.523	01:59
Air Blank		0.000	02:00
Control Test		0.076	02:00
Air Blank		0.000	02:01
Diagnostics Check		OK	02:01

Data Download Date/Time: 04/29/2019 15:18

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREIRA,JOSE,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: PEREIRA,JOSE
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: CHUB AC

DATE/TIME OF TEST: 04/14/2019 05:17:24

Diagnostics Check		OK	05:11
Air Blank		0.000	05:11
Control Test		0.081	05:11
Air Blank		0.000	05:12
Subject Sample	# 1	0.185	05:12
Breath Volume		1.203	05:12
Air Blank		0.000	05:13
Air Blank		0.000	05:15
Subject Sample	# 2	0.187	05:15
Breath Volume		1.339	05:15
Air Blank		0.000	05:16
Control Test		0.080	05:16
Air Blank		0.000	05:17
Diagnostics Check		OK	05:17

Data Download Date/Time: 04/29/2019 15:18

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREIRA,JOSE,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: PEREIRA,JOSE
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 06:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:44

SUBJECT'S LAST NAME: GRAZIANO

DATE/TIME OF TEST: 04/20/2019 07:11:23

Diagnostics Check		OK	07:04
Air Blank		0.000	07:05
Control Test		0.080	07:05
Air Blank		0.000	07:05
Subject Sample	# 1	0.088	07:06
Breath Volume		2.531	07:06
Air Blank		0.000	07:07
Air Blank		0.000	07:09
Subject Sample	# 2	0.084	07:09
Breath Volume		2.160	07:09
Air Blank		0.000	07:10
Control Test		0.080	07:10
Air Blank		0.000	07:11
Diagnostics Check		OK	07:11

Data Download Date/Time: 04/29/2019 15:18

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOLNAR,JARROD,T
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: MOLNAR,JARROD
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 14:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:30

SUBJECT'S LAST NAME: HENDRIX

DATE/TIME OF TEST: 04/20/2019 15:18:03

Diagnostics Check		OK	15:11
Air Blank		0.000	15:12
Control Test		0.080	15:12
Air Blank		0.000	15:13
Subject Sample	# 1	0.234	15:13
Breath Volume		1.667	15:13
Air Blank		0.000	15:14
Air Blank		0.000	15:16
Subject Sample	# 2	0.231	15:16
Breath Volume		1.859	15:16
Air Blank		0.000	15:17
Control Test		0.079	15:17
Air Blank		0.000	15:17
Diagnostics Check		OK	15:18

Data Download Date/Time: 04/29/2019 15:18

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUDSON,JOSHUA,K
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: ZOLLER,JOEL
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 03/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 04/01/2019 01:27:05

Diagnostics Check		OK	01:15	
Air Blank		0.000	01:16	
Control Test		0.081	01:16	
Air Blank		0.000	01:16	
Subject Sample	# 1		01:20	No Sample Provided
Breath Volume		0.000	01:20	
Air Blank		0.000	01:20	
Air Blank		0.000	01:22	
Subject Sample	# 2		01:25	No Sample Provided
Breath Volume		0.000	01:25	
Air Blank		0.000	01:26	
Control Test		0.082	01:26	
Air Blank		0.000	01:27	
Diagnostics Check		OK	01:27	

Data Download Date/Time: 04/10/2019 07:28

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VO,SEBASTIAN,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: VO,SEBASTIAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: BARLOW

DATE/TIME OF TEST: 04/05/2019 01:33:06

Diagnostics Check		OK	01:26
Air Blank		0.000	01:27
Control Test		0.080	01:27
Air Blank		0.000	01:28
Subject Sample	# 1	0.118	01:28
Breath Volume		4.019	01:28
Air Blank		0.000	01:29
Air Blank		0.000	01:31
Subject Sample	# 2	0.117	01:31
Breath Volume		4.195	01:31
Air Blank		0.000	01:32
Control Test		0.080	01:32
Air Blank		0.000	01:33
Diagnostics Check		OK	01:33

Data Download Date/Time: 04/10/2019 07:28

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,AMANDA,L
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: PARLIER,STEPHEN
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 03/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: NADOLNY

DATE/TIME OF TEST: 04/06/2019 03:07:56

Diagnostics Check		OK	02:59
Air Blank		0.000	03:00
Control Test		0.079	03:00
Air Blank		0.000	03:01
Subject Sample	# 1	0.241	03:02
Breath Volume		2.734	03:02
Air Blank		0.000	03:02
Air Blank		0.000	03:04
Subject Sample	# 2	0.244	03:06
Breath Volume		1.687	03:06
Air Blank		0.000	03:07
Control Test		0.080	03:07
Air Blank		0.000	03:07
Diagnostics Check		OK	03:07

Data Download Date/Time: 04/10/2019 07:28

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUDSON,JOSHUA,K
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: HUDSON,JOSHUA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 03/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:43

SUBJECT'S LAST NAME: BALLARD

DATE/TIME OF TEST: 04/06/2019 23:15:18

Diagnostics Check		OK	23:08
Air Blank		0.000	23:08
Control Test		0.080	23:09
Air Blank		0.000	23:09
Subject Sample	# 1	0.092	23:10
Breath Volume		1.914	23:10
Air Blank		0.000	23:11
Air Blank		0.000	23:13
Subject Sample	# 2	0.090	23:13
Breath Volume		2.312	23:13
Air Blank		0.000	23:14
Control Test		0.080	23:14
Air Blank		0.000	23:15
Diagnostics Check		OK	23:15

Data Download Date/Time: 04/10/2019 07:28

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENLEY,WILLIAM,C
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: HENLEY,WILLIAM
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 03/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: PARAZINE

DATE/TIME OF TEST: 04/06/2019 23:44:38

Diagnostics Check		OK	23:37
Air Blank		0.000	23:38
Control Test		0.080	23:38
Air Blank		0.000	23:38
Subject Sample	# 1	0.202	23:39
Breath Volume		2.144	23:39
Air Blank		0.000	23:40
Air Blank		0.000	23:42
Subject Sample	# 2	0.198	23:43
Breath Volume		1.988	23:43
Air Blank		0.000	23:43
Control Test		0.080	23:44
Air Blank		0.000	23:44
Diagnostics Check		OK	23:44

Data Download Date/Time: 04/10/2019 07:28

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEE,JAMES,E
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: LEE,JAMES
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 03/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 04/07/2019 01:12:29

Diagnostics Check		OK	01:06
Air Blank		0.000	01:06
Control Test		0.080	01:06
Air Blank		0.000	01:07
Subject Sample	# 1	0.122	01:07
Breath Volume		2.875	01:07
Air Blank		0.000	01:08
Air Blank		0.000	01:10
Subject Sample	# 2	0.125	01:10
Breath Volume		2.738	01:10
Air Blank		0.000	01:11
Control Test		0.081	01:11
Air Blank		0.000	01:12
Diagnostics Check		OK	01:12

Data Download Date/Time: 04/10/2019 07:28

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUDSON,JOSHUA,K
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: HUDSON,JOSHUA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 03/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: BROADWATER

DATE/TIME OF TEST: 04/07/2019 02:19:54

Diagnostics Check		OK	02:13
Air Blank		0.000	02:13
Control Test		0.080	02:14
Air Blank		0.000	02:14
Subject Sample	# 1	0.274	02:15
Breath Volume		3.902	02:15
Air Blank		0.000	02:15
Air Blank		0.000	02:17
Subject Sample	# 2	0.275	02:18
Breath Volume		4.070	02:18
Air Blank		0.000	02:18
Control Test		0.080	02:19
Air Blank		0.000	02:19
Diagnostics Check		OK	02:19

Data Download Date/Time: 04/10/2019 07:28

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUZBEE,HEATHER,N
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: PARLIER,STEVEN
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 03/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: GATES

DATE/TIME OF TEST: 04/07/2019 04:42:31

Diagnostics Check		OK	04:30	
Air Blank		0.000	04:31	
Control Test		0.079	04:31	
Air Blank		0.000	04:32	
Subject Sample	# 1	0.161	04:35	Volume Not Met
Breath Volume		0.062	04:35	
Air Blank		0.000	04:36	
Air Blank		0.000	04:37	
Subject Sample	# 2	0.146	04:40	Volume Not Met
Breath Volume		0.101	04:40	
Air Blank		0.000	04:41	
Control Test		0.079	04:41	
Air Blank		0.000	04:42	
Diagnostics Check		OK	04:42	

Data Download Date/Time: 04/10/2019 07:28

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUZBEE,HEATHER,N
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: PARLIER,STEVEN
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 03/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: GATES

DATE/TIME OF TEST: 04/07/2019 04:58:03

Diagnostics Check		OK	04:46	
Air Blank		0.000	04:46	
Control Test		0.080	04:47	
Air Blank		0.000	04:47	
Subject Sample	# 1	0.147	04:50	Volume Not Met
Breath Volume		0.101	04:50	
Air Blank		0.000	04:51	
Air Blank		0.000	04:53	
Subject Sample	# 2	0.131	04:56	Volume Not Met
Breath Volume		0.050	04:56	
Air Blank		0.000	04:57	
Control Test		0.079	04:57	
Air Blank		0.000	04:57	
Diagnostics Check		OK	04:58	

Data Download Date/Time: 04/10/2019 07:28

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEEKS,JEREMIAH,S
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: MEEKS,JEREMIAH
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 03/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 20:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:24

SUBJECT'S LAST NAME: BOYD

DATE/TIME OF TEST: 04/09/2019 21:52:30

Diagnostics Check		OK	21:44
Air Blank		0.000	21:45
Control Test		0.081	21:45
Air Blank		0.000	21:46
Subject Sample	# 1	0.000	21:46
Breath Volume		1.214	21:46
Air Blank		0.000	21:47
Air Blank		0.000	21:49
Subject Sample	# 2	0.000	21:51
Breath Volume		1.339	21:51
Air Blank		0.000	21:51
Control Test		0.080	21:51
Air Blank		0.000	21:52
Diagnostics Check		OK	21:52

Data Download Date/Time: 04/10/2019 07:28

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLEY,RASHELL,E
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: ABDELHADI,AUSTIN
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 13:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:24

SUBJECT'S LAST NAME: CARPENTER

DATE/TIME OF TEST: 04/14/2019 14:52:10

Diagnostics Check		OK	14:44
Air Blank		0.000	14:44
Control Test		0.080	14:45
Air Blank		0.000	14:45
Subject Sample	# 1	0.000	14:46
Breath Volume		1.863	14:46
Air Blank		0.000	14:47
Air Blank		0.000	14:49
Subject Sample	# 2	0.000	14:50
Breath Volume		1.175	14:50
Air Blank		0.000	14:51
Control Test		0.079	14:51
Air Blank		0.000	14:52
Diagnostics Check		OK	14:52

Data Download Date/Time: 05/06/2019 20:59

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANGELISTA,CURTIS,J
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: MALDEN,BIANCA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 17:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:24

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 04/17/2019 18:59:19

Diagnostics Check		OK	18:53
Air Blank		0.000	18:53
Control Test		0.080	18:53
Air Blank		0.000	18:54
Subject Sample	# 1	0.000	18:55
Breath Volume		1.613	18:55
Air Blank		0.000	18:55
Air Blank		0.000	18:57
Subject Sample	# 2	0.000	18:57
Breath Volume		1.855	18:57
Air Blank		0.000	18:58
Control Test		0.080	18:58
Air Blank		0.000	18:59
Diagnostics Check		OK	18:59

Data Download Date/Time: 05/06/2019 21:00

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMPTON,LOGAN,I
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: HAMPTON,LOGAN
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: LINN

DATE/TIME OF TEST: 04/17/2019 23:08:15

Diagnostics Check		OK	22:58
Air Blank		0.000	22:59
Control Test		0.080	22:59
Air Blank		0.000	22:59
Subject Sample	# 1	0.266	23:00
Breath Volume		1.476	23:00
Air Blank		0.000	23:01
Air Blank		0.000	23:03
Subject Sample	# 2	0.292	23:03
Breath Volume		2.906	23:03
Air Blank		0.000	23:04
Air Blank		0.000	23:06
Subject Sample	# 3	0.266	23:06
Breath Volume		1.878	23:06
Air Blank		0.000	23:07
Control Test		0.077	23:07
Air Blank		0.000	23:08
Diagnostics Check		OK	23:08

Data Download Date/Time: 05/06/2019 21:00

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,AMANDA,L
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: DAVIS,AMANDA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 04/19/2019 22:24:12

Diagnostics Check		OK	22:17
Air Blank		0.000	22:17
Control Test		0.080	22:18
Air Blank		0.000	22:18
Subject Sample	# 1	0.124	22:19
Breath Volume		1.382	22:19
Air Blank		0.000	22:20
Air Blank		0.000	22:21
Subject Sample	# 2	0.137	22:22
Breath Volume		2.335	22:22
Air Blank		0.000	22:23
Control Test		0.080	22:23
Air Blank		0.000	22:24
Diagnostics Check		OK	22:24

Data Download Date/Time: 05/06/2019 21:00

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,AMANDA,L
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: PARLIER,STEVEN
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: WASSERMAN

DATE/TIME OF TEST: 04/20/2019 01:02:42

Diagnostics Check		OK	00:52	
Air Blank		0.000	00:53	
Control Test		0.080	00:53	
Air Blank		0.000	00:53	
Subject Sample	# 1	0.174	00:56	Volume Not Met
Breath Volume		0.324	00:56	
Air Blank		0.000	00:57	
Air Blank		0.000	00:59	
Subject Sample	# 2		01:01	Subject Test Refused
Breath Volume		0.367	01:01	
Air Blank		0.000	01:01	
Control Test		0.077	01:02	
Air Blank		0.000	01:02	
Diagnostics Check		OK	01:02	

Data Download Date/Time: 05/06/2019 21:00

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,CODY,A
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: WALKER,CODY
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 04/20/2019 22:31:18

Diagnostics Check		OK	22:24
Air Blank		0.000	22:25
Control Test		0.080	22:25
Air Blank		0.000	22:25
Subject Sample	# 1	0.186	22:26
Breath Volume		2.023	22:26
Air Blank		0.000	22:27
Air Blank		0.000	22:29
Subject Sample	# 2	0.183	22:29
Breath Volume		2.121	22:29
Air Blank		0.000	22:30
Control Test		0.080	22:30
Air Blank		0.000	22:31
Diagnostics Check		OK	22:31

Data Download Date/Time: 05/06/2019 21:00

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAMLIS,STEFAN,R
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: RIVKIN,SCOTT
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 14:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:01

SUBJECT'S LAST NAME: ARNETT

DATE/TIME OF TEST: 04/21/2019 15:48:03

Diagnostics Check		OK	15:36	
Air Blank		0.000	15:37	
Control Test		0.080	15:37	
Air Blank		0.000	15:37	
Subject Sample	# 1		15:40	No Sample Provided
Breath Volume		0.000	15:40	
Air Blank		0.000	15:41	
Air Blank		0.000	15:43	
Subject Sample	# 2		15:46	No Sample Provided
Breath Volume		0.000	15:46	
Air Blank		0.000	15:47	
Control Test		0.079	15:47	
Air Blank		0.000	15:47	
Diagnostics Check		OK	15:48	

Data Download Date/Time: 05/06/2019 21:00

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,AMANDA,L
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: PARLIER,STEVEN
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: ELLIS

DATE/TIME OF TEST: 04/22/2019 01:08:19

Diagnostics Check		OK	01:02
Air Blank		0.000	01:02
Control Test		0.081	01:02
Air Blank		0.000	01:03
Subject Sample	# 1	0.259	01:03
Breath Volume		2.671	01:03
Air Blank		0.000	01:04
Air Blank		0.000	01:06
Subject Sample	# 2	0.260	01:06
Breath Volume		3.414	01:06
Air Blank		0.000	01:07
Control Test		0.080	01:07
Air Blank		0.000	01:08
Diagnostics Check		OK	01:08

Data Download Date/Time: 05/06/2019 21:00

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEACH,ZACHARY,M
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: LEACH,ZACHARY
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: MCLESTER

DATE/TIME OF TEST: 04/23/2019 22:46:27

Diagnostics Check		OK	22:34	
Air Blank		0.000	22:34	
Control Test		0.079	22:35	
Air Blank		0.000	22:35	
Subject Sample	# 1	0.208	22:38	Slope Not Level
Breath Volume		1.570	22:38	
Air Blank		0.000	22:39	
Air Blank		0.000	22:41	
Subject Sample	# 2	0.239	22:41	
Breath Volume		1.492	22:41	
Air Blank		0.000	22:42	
Air Blank		0.000	22:44	
Subject Sample	# 3	0.223	22:44	
Breath Volume		2.140	22:44	
Air Blank		0.000	22:45	
Control Test		0.076	22:45	
Air Blank		0.000	22:46	
Diagnostics Check		OK	22:46	

Data Download Date/Time: 05/06/2019 21:00

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEEKS,JEREMIAH,S
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: MEEKS,JEREMIAH
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 04/24/2019 03:11:15

Diagnostics Check		OK	03:05
Air Blank		0.000	03:05
Control Test		0.080	03:05
Air Blank		0.000	03:06
Subject Sample	# 1	0.036	03:06
Breath Volume		2.562	03:06
Air Blank		0.000	03:07
Air Blank		0.000	03:09
Subject Sample	# 2	0.033	03:09
Breath Volume		2.585	03:09
Air Blank		0.000	03:10
Control Test		0.079	03:10
Air Blank		0.000	03:11
Diagnostics Check		OK	03:11

Data Download Date/Time: 05/06/2019 21:00

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELLEY,SAMUEL,L
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: GALLION,PETER
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: WARREN

DATE/TIME OF TEST: 04/25/2019 22:22:17

Diagnostics Check		OK	22:15
Air Blank		0.000	22:16
Control Test		0.080	22:16
Air Blank		0.000	22:16
Subject Sample	# 1	0.183	22:17
Breath Volume		1.925	22:17
Air Blank		0.000	22:18
Air Blank		0.000	22:19
Subject Sample	# 2	0.190	22:20
Breath Volume		2.183	22:20
Air Blank		0.000	22:21
Control Test		0.079	22:21
Air Blank		0.000	22:22
Diagnostics Check		OK	22:22

Data Download Date/Time: 05/06/2019 21:00

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEARN,SHELDON,J
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: ZOLLER,JOEL
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 07:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:22

SUBJECT'S LAST NAME: DONALD

DATE/TIME OF TEST: 04/28/2019 07:56:27

Diagnostics Check		OK	07:49
Air Blank		0.000	07:49
Control Test		0.080	07:50
Air Blank		0.000	07:50
Subject Sample	# 1	0.000	07:51
Breath Volume		2.117	07:51
Air Blank		0.000	07:52
Air Blank		0.000	07:54
Subject Sample	# 2	0.000	07:55
Breath Volume		2.128	07:55
Air Blank		0.000	07:55
Control Test		0.079	07:55
Air Blank		0.000	07:56
Diagnostics Check		OK	07:56

Data Download Date/Time: 05/06/2019 21:00

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,JAMIE,L
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: NELSON,JAMIE
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: MCCRARY

DATE/TIME OF TEST: 04/28/2019 20:45:19

Diagnostics Check		OK	20:42	
Air Blank		0.000	20:42	
Control Test		0.080	20:43	
Air Blank		0.000	20:43	
Subject Sample	# 1		20:43	Subject Test Refused
Breath Volume		0.000	20:43	
Air Blank		0.000	20:44	
Control Test		0.080	20:44	
Air Blank		0.000	20:45	
Diagnostics Check		OK	20:45	

Data Download Date/Time: 05/06/2019 21:00

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENLEY,WILLIAM,C
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: HENLEY,WILLIAM
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: CORKER

DATE/TIME OF TEST: 04/29/2019 02:45:51

Diagnostics Check		OK	02:39
Air Blank		0.000	02:39
Control Test		0.081	02:40
Air Blank		0.000	02:40
Subject Sample	# 1	0.208	02:41
Breath Volume		1.660	02:41
Air Blank		0.000	02:41
Air Blank		0.000	02:43
Subject Sample	# 2	0.196	02:44
Breath Volume		1.804	02:44
Air Blank		0.000	02:44
Control Test		0.080	02:45
Air Blank		0.000	02:45
Diagnostics Check		OK	02:45

Data Download Date/Time: 05/06/2019 21:00

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMPTON,LOGAN,I
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: HAMPTON,LOGAN
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 20:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:54

SUBJECT'S LAST NAME: DE LA CRUZ GONZALEZ

DATE/TIME OF TEST: 04/29/2019 21:22:18

Diagnostics Check		OK	21:15
Air Blank		0.000	21:16
Control Test		0.079	21:16
Air Blank		0.000	21:17
Subject Sample	# 1	0.154	21:17
Breath Volume		2.386	21:17
Air Blank		0.000	21:18
Air Blank		0.000	21:20
Subject Sample	# 2	0.158	21:20
Breath Volume		2.960	21:20
Air Blank		0.000	21:21
Control Test		0.079	21:21
Air Blank		0.000	21:22
Diagnostics Check		OK	21:22

Data Download Date/Time: 05/06/2019 21:00

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER,RAYMOND,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SPENCER,RAYMOND
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805702
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: GOODGER

DATE/TIME OF TEST: 04/03/2019 02:33:28

Diagnostics Check		OK	02:26
Air Blank		0.000	02:26
Control Test		0.083	02:27
Air Blank		0.000	02:27
Subject Sample	# 1	0.206	02:28
Breath Volume		2.925	02:28
Air Blank		0.000	02:29
Air Blank		0.000	02:31
Subject Sample	# 2	0.189	02:31
Breath Volume		2.246	02:31
Air Blank		0.000	02:32
Control Test		0.081	02:32
Air Blank		0.000	02:33
Diagnostics Check		OK	02:33

Data Download Date/Time: 04/30/2019 15:23

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHOTSCH,RYAN,A
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SCHOTSCH,RYAN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805702
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: COLVIN

DATE/TIME OF TEST: 04/04/2019 04:52:06

Diagnostics Check		OK	04:49	
Air Blank		0.000	04:49	
Control Test		0.081	04:50	
Air Blank		0.000	04:50	
Subject Sample	# 1		04:50	Subject Test Refused
Breath Volume		0.000	04:50	
Air Blank		0.000	04:51	
Control Test		0.082	04:51	
Air Blank		0.000	04:52	
Diagnostics Check		OK	04:52	

Data Download Date/Time: 04/30/2019 15:23

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVIER,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: LEVIER,MATTHEW
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805702
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 18:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:18

SUBJECT'S LAST NAME: MEDLOCK

DATE/TIME OF TEST: 04/05/2019 20:00:10

Diagnostics Check		OK	19:48	
Air Blank		0.000	19:49	
Control Test		0.081	19:49	
Air Blank		0.000	19:50	
Subject Sample	# 1		19:53	No Sample Provided
Breath Volume		0.000	19:53	
Air Blank		0.000	19:53	
Air Blank		0.000	19:55	
Subject Sample	# 2		19:58	No Sample Provided
Breath Volume		0.000	19:58	
Air Blank		0.000	19:59	
Control Test		0.083	19:59	
Air Blank		0.000	20:00	
Diagnostics Check		OK	20:00	

Data Download Date/Time: 04/30/2019 15:23

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARVOTH,DANIEL,P
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: HARVOTH,DANIEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805702
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: MALTAIS

DATE/TIME OF TEST: 04/06/2019 03:31:58

Diagnostics Check		OK	03:23	
Air Blank		0.000	03:23	
Control Test		0.082	03:24	
Air Blank		0.000	03:24	
Subject Sample	# 1		03:27	No Sample Provided
Breath Volume		0.000	03:27	
Air Blank		0.000	03:28	
Air Blank		0.000	03:30	
Subject Sample	# 2		03:30	Subject Test Refused
Breath Volume		0.000	03:30	
Air Blank		0.000	03:31	
Control Test		0.082	03:31	
Air Blank		0.000	03:31	
Diagnostics Check		OK	03:31	

Data Download Date/Time: 04/30/2019 15:23

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOYE,JAMES,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MOYE,JAMES
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805702
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 04:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: CASE

DATE/TIME OF TEST: 04/06/2019 05:14:46

Diagnostics Check		OK	05:06	
Air Blank		0.000	05:06	
Control Test		0.082	05:07	
Air Blank		0.000	05:07	
Subject Sample	# 1		05:10	No Sample Provided
Breath Volume		0.000	05:10	
Air Blank		0.000	05:11	
Air Blank		0.000	05:13	
Subject Sample	# 2		05:13	Subject Test Refused
Breath Volume		0.000	05:13	
Air Blank		0.000	05:13	
Control Test		0.082	05:14	
Air Blank		0.000	05:14	
Diagnostics Check		OK	05:14	

Data Download Date/Time: 04/30/2019 15:23

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER,RAYMOND,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SPENCER,RAYMOND
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805702
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 04:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 04/07/2019 05:29:32

Diagnostics Check		OK	05:22
Air Blank		0.000	05:22
Control Test		0.081	05:22
Air Blank		0.000	05:23
Subject Sample	# 1	0.218	05:24
Breath Volume		3.179	05:24
Air Blank		0.000	05:25
Air Blank		0.000	05:27
Subject Sample	# 2	0.215	05:27
Breath Volume		2.753	05:27
Air Blank		0.000	05:28
Control Test		0.081	05:28
Air Blank		0.000	05:29
Diagnostics Check		OK	05:29

Data Download Date/Time: 04/30/2019 15:23

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSNAK,BRANDEN,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: RUSNAK,BRANDEN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805702
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 06:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:44

SUBJECT'S LAST NAME: STEPHEN

DATE/TIME OF TEST: 04/07/2019 07:11:58

Diagnostics Check		OK	07:04
Air Blank		0.000	07:05
Control Test		0.081	07:05
Air Blank		0.000	07:06
Subject Sample	# 1	0.162	07:07
Breath Volume		4.269	07:07
Air Blank		0.000	07:07
Air Blank		0.000	07:09
Subject Sample	# 2	0.158	07:10
Breath Volume		4.226	07:10
Air Blank		0.000	07:11
Control Test		0.080	07:11
Air Blank		0.000	07:11
Diagnostics Check		OK	07:11

Data Download Date/Time: 04/30/2019 15:23

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAPMAN,MAVERICK,E
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CHAPMAN,MAVERICK
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805702
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: BRAAWLEY

DATE/TIME OF TEST: 04/08/2019 00:22:46

Diagnostics Check		OK	00:10
Air Blank		0.000	00:10
Control Test		0.082	00:11
Air Blank		0.000	00:11
Subject Sample	# 1	0.316	00:14
Breath Volume		1.363	00:14
Air Blank		0.000	00:14
Air Blank		0.000	00:16
Subject Sample	# 2	0.293	00:17
Breath Volume		1.308	00:17
Air Blank		0.000	00:18
Air Blank		0.000	00:20
Subject Sample	# 3	0.311	00:21
Breath Volume		1.949	00:21
Air Blank		0.000	00:21
Control Test		0.077	00:22
Air Blank		0.000	00:22
Diagnostics Check		OK	00:22

Data Download Date/Time: 04/30/2019 15:23

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRIBB,BRIAN,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CRIBB,BRIAN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805702
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 06:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:40

SUBJECT'S LAST NAME: WEEKLEY

DATE/TIME OF TEST: 04/10/2019 08:11:23

Diagnostics Check		OK	08:05
Air Blank		0.000	08:05
Control Test		0.082	08:06
Air Blank		0.000	08:06
Subject Sample	# 1	0.000	08:07
Breath Volume		3.968	08:07
Air Blank		0.000	08:07
Air Blank		0.000	08:09
Subject Sample	# 2	0.000	08:10
Breath Volume		4.183	08:10
Air Blank		0.000	08:10
Control Test		0.081	08:10
Air Blank		0.000	08:11
Diagnostics Check		OK	08:11

Data Download Date/Time: 04/30/2019 15:23

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARGITT,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: HARGITT,JOSHUA
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805702
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: MEADOWS

DATE/TIME OF TEST: 04/13/2019 22:57:51

Diagnostics Check		OK	22:46	
Air Blank		0.000	22:46	
Control Test		0.081	22:47	
Air Blank		0.000	22:47	
Subject Sample	# 1		22:50	No Sample Provided
Breath Volume		0.000	22:50	
Air Blank		0.000	22:51	
Air Blank		0.000	22:53	
Subject Sample	# 2		22:56	No Sample Provided
Breath Volume		0.000	22:56	
Air Blank		0.000	22:56	
Control Test		0.082	22:57	
Air Blank		0.000	22:57	
Diagnostics Check		OK	22:57	

Data Download Date/Time: 04/30/2019 15:23

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEAVER,COREY,R
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: WEAVER,COREY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: AG805702
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: NEWMAN

DATE/TIME OF TEST: 04/14/2019 00:36:59

Diagnostics Check		OK	00:33	
Air Blank		0.000	00:34	
Control Test		0.081	00:34	
Air Blank		0.000	00:34	
Subject Sample	# 1		00:35	Subject Test Refused
Breath Volume		0.000	00:35	
Air Blank		0.000	00:36	
Control Test		0.082	00:36	
Air Blank		0.000	00:36	
Diagnostics Check		OK	00:36	

Data Download Date/Time: 04/30/2019 15:23

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,PAUL,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: RODRIGUEZ,PAUL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805702
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: HUI

DATE/TIME OF TEST: 04/14/2019 03:03:12

Diagnostics Check		OK	02:56
Air Blank		0.000	02:57
Control Test		0.082	02:57
Air Blank		0.000	02:58
Subject Sample	# 1	0.077	02:58
Breath Volume		3.031	02:58
Air Blank		0.000	02:59
Air Blank		0.000	03:01
Subject Sample	# 2	0.075	03:01
Breath Volume		3.113	03:01
Air Blank		0.000	03:02
Control Test		0.082	03:02
Air Blank		0.000	03:03
Diagnostics Check		OK	03:03

Data Download Date/Time: 04/30/2019 15:23

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER,RAYMOND,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SPENCER,RAYMOND
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805702
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: ARMSTRONG

DATE/TIME OF TEST: 04/14/2019 04:07:48

Diagnostics Check		OK	04:03	
Air Blank		0.000	04:03	
Control Test		0.082	04:03	
Air Blank		0.000	04:04	
Subject Sample	# 1		04:06	Subject Test Refused
Breath Volume		0.000	04:06	
Air Blank		0.000	04:06	
Control Test		0.082	04:07	
Air Blank		0.000	04:07	
Diagnostics Check		OK	04:07	

Data Download Date/Time: 04/30/2019 15:23

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEINER,ZACHARY,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: STEINER,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805702
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 08:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:30

SUBJECT'S LAST NAME: FAY

DATE/TIME OF TEST: 04/18/2019 10:02:00

Diagnostics Check		OK	09:55
Air Blank		0.000	09:56
Control Test		0.082	09:56
Air Blank		0.000	09:57
Subject Sample	# 1	0.000	09:57
Breath Volume		2.988	09:57
Air Blank		0.000	09:58
Air Blank		0.000	10:00
Subject Sample	# 2	0.000	10:00
Breath Volume		2.667	10:00
Air Blank		0.000	10:01
Control Test		0.082	10:01
Air Blank		0.000	10:01
Diagnostics Check		OK	10:02

Data Download Date/Time: 04/30/2019 15:23

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARVOTH,DANIEL,P
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: HARVOTH,DANIEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805702
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: PRICE

DATE/TIME OF TEST: 04/20/2019 04:10:37

Diagnostics Check		OK	04:06	
Air Blank		0.000	04:06	
Control Test		0.082	04:06	
Air Blank		0.000	04:07	
Subject Sample	# 1		04:09	Subject Test Refused
Breath Volume		0.000	04:09	
Air Blank		0.000	04:09	
Control Test		0.082	04:10	
Air Blank		0.000	04:10	
Diagnostics Check		OK	04:10	

Data Download Date/Time: 04/30/2019 15:23

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JOHN,F
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: YOUNG,JOHN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805702
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 12:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:06

SUBJECT'S LAST NAME: SARTIPI

DATE/TIME OF TEST: 04/20/2019 13:38:37

Diagnostics Check		OK	13:26
Air Blank		0.000	13:27
Control Test		0.081	13:27
Air Blank		0.000	13:28
Subject Sample	# 1	0.457	13:29
Breath Volume		1.109	13:29
Air Blank		0.000	13:30
Air Blank		0.000	13:32
Subject Sample	# 2	0.433	13:33
Breath Volume		1.289	13:33
Air Blank		0.000	13:34
Air Blank		0.000	13:36
Subject Sample	# 3	0.456	13:36
Breath Volume		1.152	13:36
Air Blank		0.000	13:37
Control Test		0.078	13:37
Air Blank		0.000	13:38
Diagnostics Check		OK	13:38

Data Download Date/Time: 04/30/2019 15:23

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWTON,CARSON,H
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: YOUNGHANSE,JOHN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805702
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 12:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:19

SUBJECT'S LAST NAME: COOKE

DATE/TIME OF TEST: 04/20/2019 14:40:06

Diagnostics Check		OK	14:31	
Air Blank		0.000	14:32	
Control Test		0.081	14:32	
Air Blank		0.000	14:32	
Subject Sample	# 1		14:36	No Sample Provided
Breath Volume		0.000	14:36	
Air Blank		0.000	14:36	
Air Blank		0.000	14:38	
Subject Sample	# 2		14:38	Subject Test Refused
Breath Volume		0.000	14:38	
Air Blank		0.000	14:39	
Control Test		0.081	14:39	
Air Blank		0.000	14:40	
Diagnostics Check		OK	14:40	

Data Download Date/Time: 04/30/2019 15:23

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMATEER,TRACY,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BAKER,RYAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805702
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:59

SUBJECT'S LAST NAME: NORTHRUP

DATE/TIME OF TEST: 04/20/2019 15:24:49

Diagnostics Check		OK	15:19	
Air Blank		0.000	15:19	
Control Test		0.082	15:20	
Air Blank		0.000	15:20	
Subject Sample	# 1	0.000	15:21	
Breath Volume		1.828	15:21	
Air Blank		0.000	15:21	
Air Blank			15:23	RFI Detect
Air Blank		0.000	15:23	
Control Test		0.083	15:24	
Air Blank		0.000	15:24	
Diagnostics Check		OK	15:24	

Data Download Date/Time: 04/30/2019 15:24

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEAVER,COREY,R
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: WEAVER,COREY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805702
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 04:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:58

SUBJECT'S LAST NAME: WU

DATE/TIME OF TEST: 04/21/2019 05:39:11

Diagnostics Check		OK	05:22	
Air Blank		0.000	05:23	
Control Test		0.082	05:23	
Air Blank		0.000	05:24	
Subject Sample	# 1		05:27	No Sample Provided
Breath Volume		0.000	05:27	
Air Blank		0.000	05:27	
Air Blank		0.000	05:29	
Subject Sample	# 2	0.093	05:31	
Breath Volume		1.375	05:31	
Air Blank		0.000	05:32	
Air Blank		0.000	05:34	
Subject Sample	# 3	0.089	05:37	Volume Not Met
Breath Volume		0.800	05:37	
Air Blank		0.000	05:38	
Control Test		0.080	05:38	
Air Blank		0.000	05:39	
Diagnostics Check		OK	05:39	

Data Download Date/Time: 04/30/2019 15:24

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOZARD,CHRYSTAL,D
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: BOZARD,CHRYSTAL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805702
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 17:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:58

SUBJECT'S LAST NAME: MYERS

DATE/TIME OF TEST: 04/25/2019 18:29:54

Diagnostics Check		OK	18:18	
Air Blank		0.000	18:18	
Control Test		0.081	18:19	
Air Blank		0.000	18:19	
Subject Sample	# 1		18:22	No Sample Provided
Breath Volume		0.000	18:22	
Air Blank		0.000	18:23	
Air Blank		0.000	18:25	
Subject Sample	# 2		18:28	No Sample Provided
Breath Volume		0.000	18:28	
Air Blank		0.000	18:28	
Control Test		0.082	18:29	
Air Blank		0.000	18:29	
Diagnostics Check		OK	18:29	

Data Download Date/Time: 04/30/2019 15:24

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,DANIEL,K
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: DAVIS,DANIEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805702
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: WHITFIELD

DATE/TIME OF TEST: 04/26/2019 03:30:58

Diagnostics Check		OK	03:24
Air Blank		0.000	03:25
Control Test		0.082	03:25
Air Blank		0.000	03:25
Subject Sample	# 1	0.111	03:26
Breath Volume		1.308	03:26
Air Blank		0.000	03:27
Air Blank		0.000	03:28
Subject Sample	# 2	0.111	03:29
Breath Volume		1.562	03:29
Air Blank		0.000	03:30
Control Test		0.081	03:30
Air Blank		0.000	03:30
Diagnostics Check		OK	03:30

Data Download Date/Time: 04/30/2019 15:24

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAPMAN,MAVERICK,E
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CHAPMAN,MAVERICK
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805702
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: PAULEY

DATE/TIME OF TEST: 04/26/2019 04:19:01

Diagnostics Check		OK	04:13
Air Blank		0.000	04:13
Control Test		0.082	04:13
Air Blank		0.000	04:14
Subject Sample	# 1	0.000	04:14
Breath Volume		4.003	04:14
Air Blank		0.000	04:15
Air Blank		0.000	04:17
Subject Sample	# 2	0.000	04:17
Breath Volume		3.722	04:17
Air Blank		0.000	04:18
Control Test		0.081	04:18
Air Blank		0.000	04:18
Diagnostics Check		OK	04:19

Data Download Date/Time: 04/30/2019 15:24

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAINE,JACK,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BLAINE,JACK
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805702
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 15:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:54

SUBJECT'S LAST NAME: MAST

DATE/TIME OF TEST: 04/26/2019 17:18:51

Diagnostics Check		OK	17:15	
Air Blank		0.000	17:15	
Control Test		0.082	17:16	
Air Blank		0.000	17:16	
Subject Sample	# 1		17:17	Subject Test Refused
Breath Volume		0.000	17:17	
Air Blank		0.000	17:17	
Control Test		0.082	17:18	
Air Blank		0.000	17:18	
Diagnostics Check		OK	17:18	

Data Download Date/Time: 04/30/2019 15:24

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOUGH,AARON,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: GOUGH,AARON
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805702
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 04/27/2019 00:26:48

Diagnostics Check		OK	00:20
Air Blank		0.000	00:20
Control Test		0.082	00:20
Air Blank		0.000	00:21
Subject Sample	# 1	0.000	00:22
Breath Volume		2.121	00:22
Air Blank		0.000	00:22
Air Blank		0.000	00:24
Subject Sample	# 2	0.000	00:25
Breath Volume		2.257	00:25
Air Blank		0.000	00:25
Control Test		0.082	00:26
Air Blank		0.000	00:26
Diagnostics Check		OK	00:26

Data Download Date/Time: 04/30/2019 15:24

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAPMAN,MAVERICK,E
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CHAPMAN,MAVERICK
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805702
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: PYRITZ

DATE/TIME OF TEST: 04/27/2019 03:14:33

Diagnostics Check		OK	03:08
Air Blank		0.000	03:08
Control Test		0.082	03:09
Air Blank		0.000	03:09
Subject Sample	# 1	0.182	03:10
Breath Volume		3.035	03:10
Air Blank		0.000	03:10
Air Blank		0.000	03:12
Subject Sample	# 2	0.182	03:13
Breath Volume		3.273	03:13
Air Blank		0.000	03:13
Control Test		0.081	03:13
Air Blank		0.000	03:14
Diagnostics Check		OK	03:14

Data Download Date/Time: 04/30/2019 15:24

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHOTSCH,RYAN,A
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SCHOTSCH,RYAN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805702
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 04:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: GWAIKOLO

DATE/TIME OF TEST: 04/27/2019 05:20:16

Diagnostics Check		OK	05:11
Air Blank		0.000	05:12
Control Test		0.082	05:12
Air Blank		0.000	05:13
Subject Sample	# 1	0.115	05:15
Breath Volume		1.902	05:15
Air Blank		0.000	05:16
Air Blank		0.000	05:17
Subject Sample	# 2	0.119	05:18
Breath Volume		2.242	05:18
Air Blank		0.000	05:19
Control Test		0.082	05:19
Air Blank		0.000	05:20
Diagnostics Check		OK	05:20

Data Download Date/Time: 04/30/2019 15:24

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,JILLIAN,R
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: FERNANDEZ,JILLIAN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805702
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:36

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 04/28/2019 05:04:34

Diagnostics Check		OK	04:57
Air Blank		0.000	04:57
Control Test		0.082	04:58
Air Blank		0.000	04:58
Subject Sample	# 1	0.149	05:00
Breath Volume		2.453	05:00
Air Blank		0.000	05:00
Air Blank		0.000	05:02
Subject Sample	# 2	0.148	05:02
Breath Volume		1.781	05:02
Air Blank		0.000	05:03
Control Test		0.082	05:03
Air Blank		0.000	05:04
Diagnostics Check		OK	05:04

Data Download Date/Time: 04/30/2019 15:24

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEINER,ZACHARY,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: STEINER,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805702
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 17:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: BALENTINE

DATE/TIME OF TEST: 04/28/2019 18:24:34

Diagnostics Check		OK	18:21	
Air Blank		0.000	18:21	
Control Test		0.082	18:22	
Air Blank		0.000	18:22	
Subject Sample	# 1		18:23	Subject Test Refused
Breath Volume		0.000	18:23	
Air Blank		0.000	18:23	
Control Test		0.083	18:23	
Air Blank		0.000	18:24	
Diagnostics Check		OK	18:24	

Data Download Date/Time: 04/30/2019 15:24

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SINGLETARY,MATHEW,C
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SINGLETARY,MATHEW
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 03/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 16:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: STARLING

DATE/TIME OF TEST: 04/04/2019 17:26:12

Diagnostics Check		OK	17:20
Air Blank		0.000	17:20
Control Test		0.081	17:21
Air Blank		0.000	17:21
Subject Sample	# 1	0.000	17:21
Breath Volume		1.316	17:21
Air Blank		0.000	17:22
Air Blank		0.000	17:24
Subject Sample	# 2	0.000	17:24
Breath Volume		1.359	17:24
Air Blank		0.000	17:25
Control Test		0.079	17:25
Air Blank		0.000	17:26
Diagnostics Check		OK	17:26

Data Download Date/Time: 04/27/2019 07:10

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUSGROVE,DUSTIN,K
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MUSGROVE,DUSTIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 03/24/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 22:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 04/04/2019 22:58:44

Diagnostics Check	OK	22:57	
Air Blank	0.000	22:58	
Control Test	0.000	22:58	RFI Detect
Air Blank	0.000	22:58	

Data Download Date/Time: 04/27/2019 07:10

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUSGROVE,DUSTIN,K
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MUSGROVE,DUSTIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 03/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 22:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 04/04/2019 23:06:47

Diagnostics Check		OK	23:00
Air Blank		0.000	23:00
Control Test		0.081	23:01
Air Blank		0.000	23:01
Subject Sample	# 1	0.149	23:02
Breath Volume		3.390	23:02
Air Blank		0.000	23:02
Air Blank		0.000	23:04
Subject Sample	# 2	0.141	23:05
Breath Volume		3.789	23:05
Air Blank		0.000	23:05
Control Test		0.079	23:06
Air Blank		0.000	23:06
Diagnostics Check		OK	23:06

Data Download Date/Time: 04/27/2019 07:10

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BALDREE,TASHAHNDA,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BALDREE,TASHAHNDA
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 03/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 06:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:40

SUBJECT'S LAST NAME: CHERRY

DATE/TIME OF TEST: 04/05/2019 07:07:22

Diagnostics Check		OK	07:00
Air Blank		0.000	07:01
Control Test		0.081	07:01
Air Blank		0.000	07:01
Subject Sample	# 1	0.116	07:02
Breath Volume		2.785	07:02
Air Blank		0.000	07:03
Air Blank		0.000	07:05
Subject Sample	# 2	0.111	07:05
Breath Volume		3.730	07:05
Air Blank		0.000	07:06
Control Test		0.077	07:06
Air Blank		0.000	07:07
Diagnostics Check		OK	07:07

Data Download Date/Time: 04/27/2019 07:10

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISTRUNK,TRENT,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SISTRUNK,TRENT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 03/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 13:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:35

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 04/07/2019 15:03:07

Diagnostics Check		OK	14:57
Air Blank		0.000	14:57
Control Test		0.080	14:58
Air Blank		0.000	14:58
Subject Sample	# 1	0.000	14:58
Breath Volume		2.554	14:58
Air Blank		0.000	14:59
Air Blank		0.000	15:01
Subject Sample	# 2	0.000	15:01
Breath Volume		2.195	15:01
Air Blank		0.000	15:02
Control Test		0.079	15:02
Air Blank		0.000	15:03
Diagnostics Check		OK	15:03

Data Download Date/Time: 04/27/2019 07:10

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CREWS,MICHAEL,B
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: COLBERT,BRANDON
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 03/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: SPICER

DATE/TIME OF TEST: 04/09/2019 01:47:07

Diagnostics Check		OK	01:40
Air Blank		0.000	01:41
Control Test		0.081	01:41
Air Blank		0.000	01:42
Subject Sample	# 1	0.060	01:42
Breath Volume		1.945	01:42
Air Blank		0.000	01:43
Air Blank		0.000	01:45
Subject Sample	# 2	0.059	01:45
Breath Volume		2.332	01:45
Air Blank		0.000	01:46
Control Test		0.080	01:46
Air Blank		0.000	01:47
Diagnostics Check		OK	01:47

Data Download Date/Time: 04/27/2019 07:10

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CREWS,MICHAEL,B
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: CREWS,MICHAEL
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 03/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 18:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:18

SUBJECT'S LAST NAME: COPELAND

DATE/TIME OF TEST: 04/12/2019 18:47:06

Diagnostics Check		OK	18:38	
Air Blank		0.000	18:39	
Control Test		0.080	18:39	
Air Blank		0.000	18:39	
Subject Sample	# 1		18:43	No Sample Provided
Breath Volume		0.000	18:43	
Air Blank		0.000	18:43	
Air Blank		0.000	18:45	
Subject Sample	# 2		18:45	Subject Test Refused
Breath Volume		0.000	18:45	
Air Blank		0.000	18:46	
Control Test		0.080	18:46	
Air Blank		0.000	18:47	
Diagnostics Check		OK	18:47	

Data Download Date/Time: 04/27/2019 07:10

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BALDREE,TASHAHNDA,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: HERNANDEZ,ALEXIS
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 03/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 18:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: CLARY

DATE/TIME OF TEST: 04/12/2019 20:32:07

Diagnostics Check	OK	20:30	
Air Blank	0.000	20:31	
Control Test	0.081	20:31	
Air Blank		20:31	RFI Detect
Air Blank	0.000	20:32	

Data Download Date/Time: 04/27/2019 07:10

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BALDREE,TASHAHNDA,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: HERNANDEZ,ALEXIS
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 03/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 18:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: CLARY

DATE/TIME OF TEST: 04/12/2019 20:40:35

Diagnostics Check		OK	20:34
Air Blank		0.000	20:34
Control Test		0.080	20:34
Air Blank		0.000	20:35
Subject Sample	# 1	0.171	20:36
Breath Volume		4.675	20:36
Air Blank		0.000	20:36
Air Blank		0.000	20:38
Subject Sample	# 2	0.153	20:38
Breath Volume		2.136	20:38
Air Blank		0.000	20:39
Control Test		0.078	20:39
Air Blank		0.000	20:40
Diagnostics Check		OK	20:40

Data Download Date/Time: 04/27/2019 07:10

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FISCHER,MICHAEL,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: FISCHER,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 03/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 13:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:25

SUBJECT'S LAST NAME: TAUNTON

DATE/TIME OF TEST: 04/14/2019 15:53:03

Diagnostics Check		OK	15:46
Air Blank		0.000	15:47
Control Test		0.079	15:47
Air Blank		0.000	15:48
Subject Sample	# 1	0.000	15:48
Breath Volume		3.898	15:48
Air Blank		0.000	15:49
Air Blank		0.000	15:51
Subject Sample	# 2	0.000	15:51
Breath Volume		4.085	15:51
Air Blank		0.000	15:52
Control Test		0.078	15:52
Air Blank		0.000	15:52
Diagnostics Check		OK	15:53

Data Download Date/Time: 04/27/2019 07:10

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CREWS,MICHAEL,B
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: MONTEMURNO,MATT
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 03/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: HAASE

DATE/TIME OF TEST: 04/19/2019 01:09:05

Diagnostics Check		OK	01:02
Air Blank		0.000	01:03
Control Test		0.080	01:03
Air Blank		0.000	01:04
Subject Sample	# 1	0.148	01:04
Breath Volume		3.671	01:04
Air Blank		0.000	01:05
Air Blank		0.000	01:07
Subject Sample	# 2	0.152	01:07
Breath Volume		3.644	01:07
Air Blank		0.000	01:08
Control Test		0.080	01:08
Air Blank		0.000	01:08
Diagnostics Check		OK	01:09

Data Download Date/Time: 04/27/2019 07:10

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAURES,BRANDYN,K
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BAURES,BRANDYN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 03/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 20:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: MILLARES

DATE/TIME OF TEST: 04/20/2019 22:07:38

Diagnostics Check		OK	22:01
Air Blank		0.000	22:01
Control Test		0.081	22:02
Air Blank		0.000	22:02
Subject Sample	# 1	0.213	22:03
Breath Volume		2.339	22:03
Air Blank		0.000	22:03
Air Blank		0.000	22:05
Subject Sample	# 2	0.211	22:06
Breath Volume		1.109	22:06
Air Blank		0.000	22:06
Control Test		0.081	22:07
Air Blank		0.000	22:07
Diagnostics Check		OK	22:07

Data Download Date/Time: 04/27/2019 07:10

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000779

INSTRUMENT REGISTERED TO: MADISON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAMES,JONATHAN,B
BREATH TEST OPERATOR AGENCY: Madison County SO
ARRESTING OFFICER: GILBERT,SHANE
ARRESTING OFFICER AGENCY: Madison County SO
LAST AGENCY INSPECTION: 04/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A3
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: VILLEREAL

DATE/TIME OF TEST: 04/06/2019 01:50:06

Diagnostics Check		OK	01:42	
Air Blank		0.000	01:42	
Control Test		0.080	01:43	
Air Blank		0.000	01:43	
Subject Sample	# 1	0.184	01:45	
Breath Volume		1.214	01:45	
Air Blank		0.000	01:46	
Air Blank		0.000	01:48	
Subject Sample	# 2	0.190	01:48	
Breath Volume		1.632	01:48	
Air Blank		0.000	01:49	
Control Test		0.073	01:49	Control Outside Tolerance
Air Blank		0.000	01:50	

Data Download Date/Time: 05/15/2019 10:17

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANKS,SEAN,C
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: BANKS,SEAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 04/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 04:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: COLE

DATE/TIME OF TEST: 04/21/2019 05:38:08

Diagnostics Check		OK	05:35	
Air Blank		0.000	05:35	
Control Test		0.080	05:36	
Air Blank		0.000	05:36	
Subject Sample	# 1		05:36	Subject Test Refused
Breath Volume		0.000	05:36	
Air Blank		0.000	05:37	
Control Test		0.080	05:37	
Air Blank		0.000	05:38	
Diagnostics Check		OK	05:38	

Data Download Date/Time: 05/14/2019 11:02

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NGUYEN,TAI,T
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: NGUYEN,TAI
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 04/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: CHAPMAN

DATE/TIME OF TEST: 04/23/2019 23:44:23

Diagnostics Check		OK	23:38
Air Blank		0.000	23:38
Control Test		0.079	23:38
Air Blank		0.000	23:39
Subject Sample	# 1	0.215	23:39
Breath Volume		2.722	23:39
Air Blank		0.000	23:40
Air Blank		0.000	23:42
Subject Sample	# 2	0.217	23:42
Breath Volume		2.492	23:42
Air Blank		0.000	23:43
Control Test		0.078	23:43
Air Blank		0.000	23:44
Diagnostics Check		OK	23:44

Data Download Date/Time: 05/14/2019 11:02

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANKS,SEAN,C
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: BANKS,SEAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 04/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 04/26/2019 00:53:25

Diagnostics Check		OK	00:50	
Air Blank		0.000	00:51	
Control Test		0.079	00:51	
Air Blank		0.000	00:51	
Subject Sample	# 1		00:52	Subject Test Refused
Breath Volume		0.000	00:52	
Air Blank		0.000	00:52	
Control Test		0.079	00:52	
Air Blank		0.000	00:53	
Diagnostics Check		OK	00:53	

Data Download Date/Time: 05/14/2019 11:02

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORONA CANCHOLA,GERARDO,
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: SCAPIN,BRANDON
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 04/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 16:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:44

SUBJECT'S LAST NAME: CORONA CANCHOLA

DATE/TIME OF TEST: 04/27/2019 17:08:59

Diagnostics Check		OK	17:04	
Air Blank		0.000	17:05	
Control Test		0.079	17:05	
Air Blank		0.000	17:05	
Subject Sample	# 1	0.194	17:07	
Breath Volume		1.554	17:07	
Air Blank			17:08	Purge Fail
Air Blank		0.000	17:08	

Data Download Date/Time: 05/14/2019 11:02

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCAPIN,BRANDON,L
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: SCAPIN,BRANDON
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 04/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 16:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:44

SUBJECT'S LAST NAME: CORONA CANCHOLA

DATE/TIME OF TEST: 04/27/2019 17:22:18

Diagnostics Check		OK	17:14
Air Blank		0.000	17:14
Control Test		0.077	17:15
Air Blank		0.000	17:15
Subject Sample	# 1	0.199	17:17
Breath Volume		1.730	17:17
Air Blank		0.000	17:17
Air Blank		0.000	17:19
Subject Sample	# 2	0.193	17:20
Breath Volume		1.460	17:20
Air Blank		0.000	17:21
Control Test		0.079	17:21
Air Blank		0.000	17:22
Diagnostics Check		OK	17:22

Data Download Date/Time: 05/14/2019 11:03

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: PATTERSON,KYLE
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: BUI
ARREST TIME: 19:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: RHEIN

DATE/TIME OF TEST: 04/13/2019 19:43:36

Diagnostics Check		OK	19:40	
Air Blank		0.000	19:41	
Control Test		0.079	19:41	
Air Blank		0.000	19:42	
Subject Sample	# 1		19:42	Subject Test Refused
Breath Volume		0.000	19:42	
Air Blank		0.000	19:42	
Control Test		0.080	19:42	
Air Blank		0.000	19:43	
Diagnostics Check		OK	19:43	

Data Download Date/Time: 04/25/2019 15:11

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: PAYNE,JAMES
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: BUI
ARREST TIME: 17:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:33

SUBJECT'S LAST NAME: KELSEY

DATE/TIME OF TEST: 04/14/2019 18:12:11

Diagnostics Check		OK	18:05
Air Blank		0.000	18:06
Control Test		0.080	18:06
Air Blank		0.000	18:07
Subject Sample	# 1	0.000	18:07
Breath Volume		3.496	18:07
Air Blank		0.000	18:08
Air Blank		0.000	18:10
Subject Sample	# 2	0.000	18:10
Breath Volume		2.628	18:10
Air Blank		0.000	18:11
Control Test		0.081	18:11
Air Blank		0.000	18:12
Diagnostics Check		OK	18:12

Data Download Date/Time: 04/25/2019 15:11

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON,KYLE,L
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: PATTERSON,KYLE
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: BUI
ARREST TIME: 19:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: WARE

DATE/TIME OF TEST: 04/14/2019 19:31:43

Diagnostics Check		OK	19:28	
Air Blank		0.000	19:29	
Control Test		0.077	19:29	
Air Blank		0.000	19:30	
Subject Sample	# 1		19:30	Subject Test Refused
Breath Volume		0.000	19:30	
Air Blank		0.000	19:30	
Control Test		0.082	19:31	
Air Blank		0.000	19:31	
Diagnostics Check		OK	19:31	

Data Download Date/Time: 04/25/2019 15:11

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HETZER-HARRIS,DONTE,S
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: KANE,SEAN
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: CALL

DATE/TIME OF TEST: 04/20/2019 20:23:45

Diagnostics Check		OK	20:15
Air Blank		0.000	20:15
Control Test		0.079	20:16
Air Blank		0.000	20:16
Subject Sample	# 1	0.198	20:18
Breath Volume		1.367	20:18
Air Blank		0.000	20:19
Air Blank		0.000	20:20
Subject Sample	# 2	0.189	20:22
Breath Volume		2.507	20:22
Air Blank		0.000	20:22
Control Test		0.075	20:23
Air Blank		0.000	20:23
Diagnostics Check		OK	20:23

Data Download Date/Time: 04/25/2019 15:11

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HETZER-HARRIS,DONTE,S
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: HETZER-HARRIS,DONTE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: SIBERON

DATE/TIME OF TEST: 04/21/2019 01:49:53

Diagnostics Check	OK	01:48	
Air Blank	0.000	01:48	
Control Test	0.081	01:49	
Air Blank		01:49	RFI Detect
Air Blank	0.000	01:49	

Data Download Date/Time: 04/25/2019 15:11

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HETZER-HARRIS,DONTE,S
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: HETZER-HARRIS,DONTE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: SIBERON

DATE/TIME OF TEST: 04/21/2019 01:59:49

Diagnostics Check		OK	01:51
Air Blank		0.000	01:52
Control Test		0.080	01:52
Air Blank		0.000	01:53
Subject Sample	# 1	0.057	01:55
Breath Volume		1.617	01:55
Air Blank		0.000	01:55
Air Blank		0.000	01:57
Subject Sample	# 2	0.053	01:58
Breath Volume		1.570	01:58
Air Blank		0.000	01:58
Control Test		0.078	01:59
Air Blank		0.000	01:59
Diagnostics Check		OK	01:59

Data Download Date/Time: 04/25/2019 15:11

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FENTON,WILLIAM,B
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: FENTON,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 13:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: WHEELER

DATE/TIME OF TEST: 04/23/2019 02:16:41

Diagnostics Check		OK	02:10
Air Blank		0.000	02:10
Control Test		0.080	02:10
Air Blank		0.000	02:11
Subject Sample	# 1	0.218	02:11
Breath Volume		2.386	02:11
Air Blank		0.000	02:12
Air Blank		0.000	02:14
Subject Sample	# 2	0.210	02:15
Breath Volume		1.945	02:15
Air Blank		0.000	02:15
Control Test		0.080	02:16
Air Blank		0.000	02:16
Diagnostics Check		OK	02:16

Data Download Date/Time: 04/25/2019 15:12

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEER,BRADLEY,J
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: MEER,BRADLEY
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 11:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:45

SUBJECT'S LAST NAME: ALBURY

DATE/TIME OF TEST: 04/24/2019 12:17:14

Diagnostics Check		OK	12:05	
Air Blank		0.000	12:05	
Control Test		0.080	12:05	
Air Blank		0.000	12:06	
Subject Sample	# 1	0.000	12:09	Volume Not Met
Breath Volume		0.125	12:09	
Air Blank		0.000	12:10	
Air Blank		0.000	12:12	
Subject Sample	# 2	0.000	12:12	
Breath Volume		1.574	12:12	
Air Blank		0.000	12:13	
Air Blank		0.000	12:15	
Subject Sample	# 3	0.000	12:15	
Breath Volume		1.425	12:15	
Air Blank		0.000	12:16	
Control Test		0.079	12:16	
Air Blank		0.000	12:17	
Diagnostics Check		OK	12:17	

Data Download Date/Time: 04/25/2019 15:12

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: SCUOPPO,ANTHONY
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 04/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: HASTINGS

DATE/TIME OF TEST: 04/26/2019 03:55:54

Diagnostics Check		OK	03:48
Air Blank		0.000	03:49
Control Test		0.078	03:49
Air Blank		0.000	03:50
Subject Sample	# 1	0.000	03:51
Breath Volume		2.007	03:51
Air Blank		0.000	03:52
Air Blank		0.000	03:54
Subject Sample	# 2	0.000	03:54
Breath Volume		1.578	03:54
Air Blank		0.000	03:54
Control Test		0.077	03:55
Air Blank		0.000	03:55
Diagnostics Check		OK	03:55

Data Download Date/Time: 05/22/2019 07:13

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: MCINERNEY,JESSE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 04/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: HAVEKOST

DATE/TIME OF TEST: 04/26/2019 04:20:41

Diagnostics Check		OK	04:13
Air Blank		0.000	04:14
Control Test		0.077	04:14
Air Blank		0.000	04:14
Subject Sample	# 1	0.189	04:15
Breath Volume		2.156	04:15
Air Blank		0.000	04:16
Air Blank		0.000	04:18
Subject Sample	# 2	0.198	04:19
Breath Volume		1.441	04:19
Air Blank		0.000	04:19
Control Test		0.077	04:20
Air Blank		0.000	04:20
Diagnostics Check		OK	04:20

Data Download Date/Time: 05/22/2019 07:13

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: MONDRAGON,LIONEL
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 04/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 04/27/2019 02:16:07

Diagnostics Check		OK	02:09
Air Blank		0.000	02:10
Control Test		0.078	02:10
Air Blank		0.000	02:11
Subject Sample	# 1	0.054	02:11
Breath Volume		2.710	02:11
Air Blank		0.000	02:12
Air Blank		0.000	02:14
Subject Sample	# 2	0.053	02:14
Breath Volume		2.308	02:14
Air Blank		0.000	02:15
Control Test		0.077	02:15
Air Blank		0.000	02:16
Diagnostics Check		OK	02:16

Data Download Date/Time: 05/22/2019 07:13

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: MONDRAGON,LIONEL
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 04/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: APARICIO MARTINEZ

DATE/TIME OF TEST: 04/28/2019 05:37:41

Diagnostics Check		OK	05:31
Air Blank		0.000	05:31
Control Test		0.078	05:31
Air Blank		0.000	05:32
Subject Sample	# 1	0.180	05:33
Breath Volume		1.429	05:33
Air Blank		0.000	05:33
Air Blank		0.000	05:35
Subject Sample	# 2	0.169	05:36
Breath Volume		1.539	05:36
Air Blank		0.000	05:36
Control Test		0.077	05:37
Air Blank		0.000	05:37
Diagnostics Check		OK	05:37

Data Download Date/Time: 05/22/2019 07:13

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: DRAWDY,LAINE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:27

SUBJECT'S LAST NAME: CASON

DATE/TIME OF TEST: 04/03/2019 22:03:57

Diagnostics Check		OK	21:57
Air Blank		0.000	21:58
Control Test		0.081	21:58
Air Blank		0.000	21:59
Subject Sample	# 1	0.045	21:59
Breath Volume		2.785	21:59
Air Blank		0.000	22:00
Air Blank		0.000	22:01
Subject Sample	# 2	0.043	22:02
Breath Volume		2.625	22:02
Air Blank		0.000	22:03
Control Test		0.081	22:03
Air Blank		0.000	22:03
Diagnostics Check		OK	22:03

Data Download Date/Time: 04/25/2019 15:09

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DALIA,DANIEL,
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: DALIA,DANIEL
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: FLYNN-KELLY

DATE/TIME OF TEST: 04/07/2019 04:03:01

Diagnostics Check		OK	03:56
Air Blank		0.000	03:56
Control Test		0.080	03:56
Air Blank		0.000	03:57
Subject Sample	# 1	0.135	03:57
Breath Volume		1.339	03:57
Air Blank		0.000	03:58
Air Blank		0.000	04:00
Subject Sample	# 2	0.144	04:01
Breath Volume		1.410	04:01
Air Blank		0.000	04:02
Control Test		0.081	04:02
Air Blank		0.000	04:02
Diagnostics Check		OK	04:03

Data Download Date/Time: 04/25/2019 15:09

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON,KYLE,L
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: PATTERSON,KYLE
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: BUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: FITZPATRICK

DATE/TIME OF TEST: 04/07/2019 20:07:40

Diagnostics Check		OK	20:01
Air Blank		0.000	20:02
Control Test		0.081	20:02
Air Blank		0.000	20:02
Subject Sample	# 1	0.144	20:03
Breath Volume		1.878	20:03
Air Blank		0.000	20:03
Air Blank		0.000	20:05
Subject Sample	# 2	0.141	20:06
Breath Volume		3.214	20:06
Air Blank		0.000	20:06
Control Test		0.081	20:07
Air Blank		0.000	20:07
Diagnostics Check		OK	20:07

Data Download Date/Time: 04/25/2019 15:09

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON,KYLE,L
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: ARNOLD,JON
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 04/07/2019 20:54:18

Diagnostics Check		OK	20:48
Air Blank		0.000	20:48
Control Test		0.081	20:48
Air Blank		0.000	20:49
Subject Sample	# 1	0.150	20:49
Breath Volume		1.949	20:49
Air Blank		0.000	20:50
Air Blank		0.000	20:52
Subject Sample	# 2	0.163	20:52
Breath Volume		1.914	20:52
Air Blank		0.000	20:53
Control Test		0.080	20:53
Air Blank		0.000	20:54
Diagnostics Check		OK	20:54

Data Download Date/Time: 04/25/2019 15:10

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: SCUOPPO,ANTHONY
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: GLOTZER

DATE/TIME OF TEST: 04/08/2019 00:22:08

Diagnostics Check		OK	00:16
Air Blank		0.000	00:16
Control Test		0.081	00:16
Air Blank		0.000	00:17
Subject Sample	# 1	0.159	00:17
Breath Volume		2.847	00:17
Air Blank		0.000	00:18
Air Blank		0.000	00:20
Subject Sample	# 2	0.154	00:20
Breath Volume		2.488	00:20
Air Blank		0.000	00:21
Control Test		0.080	00:21
Air Blank		0.000	00:22
Diagnostics Check		OK	00:22

Data Download Date/Time: 04/25/2019 15:10

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: DRAWDY,LAINE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: WEBB

DATE/TIME OF TEST: 04/08/2019 23:36:59

Diagnostics Check		OK	23:30
Air Blank		0.000	23:31
Control Test		0.081	23:31
Air Blank		0.000	23:32
Subject Sample	# 1	0.009	23:32
Breath Volume		2.277	23:32
Air Blank		0.000	23:33
Air Blank		0.000	23:35
Subject Sample	# 2	0.008	23:35
Breath Volume		1.781	23:35
Air Blank		0.000	23:36
Control Test		0.081	23:36
Air Blank		0.000	23:36
Diagnostics Check		OK	23:36

Data Download Date/Time: 04/25/2019 15:10

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,RICHARD,D
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: RIDLE,EVAN
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: MANUEL HERNANDEZ

DATE/TIME OF TEST: 04/17/2019 20:45:08

Diagnostics Check		OK	20:37
Air Blank		0.000	20:38
Control Test		0.080	20:38
Air Blank		0.000	20:38
Subject Sample	# 1	0.239	20:40
Breath Volume		2.867	20:40
Air Blank		0.000	20:41
Air Blank		0.000	20:42
Subject Sample	# 2	0.230	20:43
Breath Volume		3.027	20:43
Air Blank		0.000	20:44
Control Test		0.079	20:44
Air Blank		0.000	20:45
Diagnostics Check		OK	20:45

Data Download Date/Time: 04/25/2019 15:10

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000795

INSTRUMENT REGISTERED TO: GULF COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DICKEY,WILLIAM,Q
BREATH TEST OPERATOR AGENCY: Port St. Joe PD
ARRESTING OFFICER: DICKEY,WILLIAM
ARRESTING OFFICER AGENCY: Port St. Joe PD
LAST AGENCY INSPECTION: 03/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: VARONA

DATE/TIME OF TEST: 04/02/2019 02:42:42

Diagnostics Check		OK	02:33
Air Blank		0.000	02:34
Control Test		0.084	02:34
Air Blank		0.000	02:34
Subject Sample	# 1	0.174	02:38
Breath Volume		1.742	02:38
Air Blank		0.000	02:38
Air Blank		0.000	02:40
Subject Sample	# 2	0.154	02:41
Breath Volume		1.476	02:41
Air Blank		0.000	02:41
Control Test		0.084	02:42
Air Blank		0.000	02:42
Diagnostics Check		OK	02:42

Data Download Date/Time: 05/02/2019 12:07

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000795

INSTRUMENT REGISTERED TO: GULF COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISK,CARLOS,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: SISK,CARLOS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 11:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: MUSSER

DATE/TIME OF TEST: 04/04/2019 01:37:46

Diagnostics Check		OK	01:34	
Air Blank		0.000	01:34	
Control Test		0.084	01:35	
Air Blank		0.000	01:35	
Subject Sample	# 1		01:36	Subject Test Refused
Breath Volume		0.000	01:36	
Air Blank		0.000	01:36	
Control Test		0.085	01:37	
Air Blank		0.000	01:37	
Diagnostics Check		OK	01:37	

Data Download Date/Time: 05/02/2019 12:07

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000797

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BASS,DANIEL,K
BREATH TEST OPERATOR AGENCY: FL Dept of AG
ARRESTING OFFICER: MILTON,CARLTON
ARRESTING OFFICER AGENCY: FL Dept of AG
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 04/12/2019 03:12:01

Diagnostics Check		OK	03:05
Air Blank		0.000	03:06
Control Test		0.081	03:06
Air Blank		0.000	03:07
Subject Sample	# 1	0.067	03:07
Breath Volume		3.714	03:07
Air Blank		0.000	03:08
Air Blank		0.000	03:10
Subject Sample	# 2	0.058	03:10
Breath Volume		3.660	03:10
Air Blank		0.000	03:11
Control Test		0.081	03:11
Air Blank		0.000	03:11
Diagnostics Check		OK	03:12

Data Download Date/Time: 05/08/2019 20:33

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000797

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHELAN,MATTHEW,S
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SISTRUNK,TRENT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 10:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:38

SUBJECT'S LAST NAME: JONES-JOHNSON

DATE/TIME OF TEST: 04/15/2019 11:06:51

Diagnostics Check		OK	11:00
Air Blank		0.000	11:00
Control Test		0.082	11:00
Air Blank		0.000	11:01
Subject Sample	# 1	0.244	11:02
Breath Volume		3.140	11:02
Air Blank		0.000	11:03
Air Blank		0.000	11:04
Subject Sample	# 2	0.239	11:05
Breath Volume		1.976	11:05
Air Blank		0.000	11:05
Control Test		0.083	11:06
Air Blank		0.000	11:06
Diagnostics Check		OK	11:06

Data Download Date/Time: 05/08/2019 20:33

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HOWARD,D
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 04/01/2019 03:33:37

Diagnostics Check		OK	03:26
Air Blank		0.000	03:26
Control Test		0.080	03:27
Air Blank		0.000	03:27
Subject Sample	# 1	0.187	03:28
Breath Volume		3.511	03:28
Air Blank		0.000	03:29
Air Blank		0.000	03:31
Subject Sample	# 2	0.180	03:32
Breath Volume		2.878	03:32
Air Blank		0.000	03:32
Control Test		0.079	03:33
Air Blank		0.000	03:33
Diagnostics Check		OK	03:33

Data Download Date/Time: 04/04/2019 16:58

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: AUSTIN

DATE/TIME OF TEST: 04/02/2019 04:54:37

Diagnostics Check		OK	04:44
Air Blank		0.000	04:44
Control Test		0.079	04:44
Air Blank		0.000	04:45
Subject Sample	# 1	0.215	04:48
Breath Volume		2.128	04:48
Air Blank		0.000	04:48
Air Blank		0.000	04:50
Subject Sample	# 2	0.198	04:53
Breath Volume		1.273	04:53
Air Blank		0.000	04:53
Control Test		0.078	04:54
Air Blank		0.000	04:54
Diagnostics Check		OK	04:54

Data Download Date/Time: 04/04/2019 16:58

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKIPPER,ROGER,G
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CHU,A 3378
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 13:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:41

SUBJECT'S LAST NAME: CHIARILLO

DATE/TIME OF TEST: 04/02/2019 15:13:04

Diagnostics Check		OK	15:02
Air Blank		0.000	15:03
Control Test		0.079	15:03
Air Blank		0.000	15:04
Subject Sample	# 1	0.222	15:07
Breath Volume		1.503	15:07
Air Blank		0.000	15:07
Air Blank		0.000	15:09
Subject Sample	# 2	0.235	15:11
Breath Volume		1.671	15:11
Air Blank		0.000	15:12
Control Test		0.077	15:12
Air Blank		0.000	15:12
Diagnostics Check		OK	15:13

Data Download Date/Time: 04/04/2019 16:58

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: BRUST

DATE/TIME OF TEST: 04/03/2019 00:46:33

Diagnostics Check		OK	00:40
Air Blank		0.000	00:40
Control Test		0.080	00:41
Air Blank		0.000	00:41
Subject Sample	# 1	0.132	00:42
Breath Volume		3.343	00:42
Air Blank		0.000	00:42
Air Blank		0.000	00:44
Subject Sample	# 2	0.129	00:44
Breath Volume		3.500	00:44
Air Blank		0.000	00:45
Control Test		0.079	00:45
Air Blank		0.000	00:46
Diagnostics Check		OK	00:46

Data Download Date/Time: 04/04/2019 16:58

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LYON,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: BLANCO

DATE/TIME OF TEST: 04/03/2019 02:04:37

Diagnostics Check		OK	01:56
Air Blank		0.000	01:57
Control Test		0.080	01:57
Air Blank		0.000	01:57
Subject Sample	# 1	0.116	01:59
Breath Volume		1.941	01:59
Air Blank		0.000	02:00
Air Blank		0.000	02:02
Subject Sample	# 2	0.113	02:03
Breath Volume		1.757	02:03
Air Blank		0.000	02:03
Control Test		0.079	02:04
Air Blank		0.000	02:04
Diagnostics Check		OK	02:04

Data Download Date/Time: 04/04/2019 16:58

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FISHER,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: JODWAY

DATE/TIME OF TEST: 04/03/2019 22:30:47

Diagnostics Check		OK	22:24
Air Blank		0.000	22:24
Control Test		0.080	22:25
Air Blank		0.000	22:25
Subject Sample	# 1	0.196	22:26
Breath Volume		2.488	22:26
Air Blank		0.000	22:26
Air Blank		0.000	22:28
Subject Sample	# 2	0.195	22:29
Breath Volume		2.125	22:29
Air Blank		0.000	22:29
Control Test		0.079	22:30
Air Blank		0.000	22:30
Diagnostics Check		OK	22:30

Data Download Date/Time: 04/04/2019 16:58

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES AVERY,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WOLSTENHOLM,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: ALEXANDER

DATE/TIME OF TEST: 04/04/2019 02:38:55

Diagnostics Check		OK	02:32
Air Blank		0.000	02:33
Control Test		0.080	02:33
Air Blank		0.000	02:33
Subject Sample	# 1	0.181	02:34
Breath Volume		1.292	02:34
Air Blank		0.000	02:34
Air Blank		0.000	02:36
Subject Sample	# 2	0.173	02:37
Breath Volume		1.289	02:37
Air Blank		0.000	02:38
Control Test		0.079	02:38
Air Blank		0.000	02:38
Diagnostics Check		OK	02:38

Data Download Date/Time: 04/04/2019 16:58

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKIPPER,ROGER,G
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HOSKINS,H 4425
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 09:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:35

SUBJECT'S LAST NAME: POTTER

DATE/TIME OF TEST: 04/04/2019 11:05:41

Diagnostics Check		OK	10:58
Air Blank		0.000	10:58
Control Test		0.079	10:58
Air Blank		0.000	10:59
Subject Sample	# 1	0.173	11:00
Breath Volume		1.949	11:00
Air Blank		0.000	11:01
Air Blank		0.000	11:02
Subject Sample	# 2	0.166	11:04
Breath Volume		2.273	11:04
Air Blank		0.000	11:04
Control Test		0.079	11:05
Air Blank		0.000	11:05
Diagnostics Check		OK	11:05

Data Download Date/Time: 04/04/2019 16:58

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000809

INSTRUMENT REGISTERED TO: DIXIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANUS,GLEN,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: GANUS,GLEN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1761808A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 04/29/2019 06:25:50

Diagnostics Check		OK	06:23	
Air Blank		0.000	06:23	
Control Test		0.081	06:23	
Air Blank		0.000	06:24	
Subject Sample	# 1		06:24	Subject Test Refused
Breath Volume		0.000	06:24	
Air Blank		0.000	06:24	
Control Test		0.081	06:25	
Air Blank		0.000	06:25	
Diagnostics Check		OK	06:25	

Data Download Date/Time: 04/30/2019 11:03

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,MICHAEL,T
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: THOMAS,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: LASHLEY

DATE/TIME OF TEST: 04/01/2019 01:42:53

Diagnostics Check		OK	01:36
Air Blank		0.000	01:37
Control Test		0.081	01:37
Air Blank		0.000	01:37
Subject Sample	# 1	0.253	01:38
Breath Volume		1.953	01:38
Air Blank		0.000	01:39
Air Blank		0.000	01:40
Subject Sample	# 2	0.254	01:41
Breath Volume		1.554	01:41
Air Blank		0.000	01:41
Control Test		0.080	01:42
Air Blank		0.000	01:42
Diagnostics Check		OK	01:42

Data Download Date/Time: 04/22/2019 10:04

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REID,MICHAEL,G
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: REID,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 08:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: KEPLINGER

DATE/TIME OF TEST: 04/02/2019 09:52:29

Diagnostics Check		OK	09:46
Air Blank		0.000	09:46
Control Test		0.081	09:47
Air Blank		0.000	09:47
Subject Sample	# 1	0.000	09:48
Breath Volume		2.542	09:48
Air Blank		0.000	09:48
Air Blank		0.000	09:50
Subject Sample	# 2	0.000	09:51
Breath Volume		3.113	09:51
Air Blank		0.000	09:51
Control Test		0.081	09:51
Air Blank		0.000	09:52
Diagnostics Check		OK	09:52

Data Download Date/Time: 04/22/2019 10:04

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,D,E
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: WRIGHT,D
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: MANCINI

DATE/TIME OF TEST: 04/04/2019 02:09:22

Diagnostics Check		OK	02:02
Air Blank		0.000	02:03
Control Test		0.081	02:03
Air Blank		0.000	02:04
Subject Sample	# 1	0.224	02:04
Breath Volume		2.867	02:04
Air Blank		0.000	02:05
Air Blank		0.000	02:07
Subject Sample	# 2	0.205	02:07
Breath Volume		1.425	02:07
Air Blank		0.000	02:08
Control Test		0.082	02:08
Air Blank		0.000	02:09
Diagnostics Check		OK	02:09

Data Download Date/Time: 04/22/2019 10:04

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOLLAZZO,C,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SOLLAZZO,C
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 04:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:07

SUBJECT'S LAST NAME: BLESSIE

DATE/TIME OF TEST: 04/07/2019 06:52:08

Diagnostics Check		OK	06:45
Air Blank		0.000	06:46
Control Test		0.081	06:46
Air Blank		0.000	06:47
Subject Sample	# 1	0.165	06:47
Breath Volume		1.222	06:47
Air Blank		0.000	06:48
Air Blank		0.000	06:49
Subject Sample	# 2	0.159	06:50
Breath Volume		1.660	06:50
Air Blank		0.000	06:51
Control Test		0.081	06:51
Air Blank		0.000	06:52
Diagnostics Check		OK	06:52

Data Download Date/Time: 04/22/2019 10:04

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTIGLIONE,SALVATORE,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: CASTIGLIONE,SALVATORE
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: AQUINO

DATE/TIME OF TEST: 04/08/2019 00:23:48

Diagnostics Check		OK	00:17
Air Blank		0.000	00:18
Control Test		0.080	00:18
Air Blank		0.000	00:19
Subject Sample	# 1	0.169	00:19
Breath Volume		1.886	00:19
Air Blank		0.000	00:20
Air Blank		0.000	00:21
Subject Sample	# 2	0.159	00:22
Breath Volume		1.304	00:22
Air Blank		0.000	00:22
Control Test		0.080	00:23
Air Blank		0.000	00:23
Diagnostics Check		OK	00:23

Data Download Date/Time: 04/22/2019 10:04

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPE,CALVIN,W
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: FITCH,ANGIE
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 04/08/2019 02:24:46

Diagnostics Check		OK	02:17
Air Blank		0.000	02:18
Control Test		0.080	02:18
Air Blank		0.000	02:19
Subject Sample	# 1	0.387	02:20
Breath Volume		3.839	02:20
Air Blank		0.000	02:20
Air Blank		0.000	02:22
Subject Sample	# 2	0.373	02:23
Breath Volume		4.187	02:23
Air Blank		0.000	02:23
Control Test		0.079	02:24
Air Blank		0.000	02:24
Diagnostics Check		OK	02:24

Data Download Date/Time: 04/22/2019 10:04

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCALANTE,JOSE,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: COON,AUSTIN
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: ESCALANTE

DATE/TIME OF TEST: 04/08/2019 22:22:01

Diagnostics Check		OK	22:15
Air Blank		0.000	22:15
Control Test		0.081	22:16
Air Blank		0.000	22:16
Subject Sample	# 1	0.268	22:17
Breath Volume		1.464	22:17
Air Blank		0.000	22:17
Air Blank		0.000	22:19
Subject Sample	# 2	0.263	22:20
Breath Volume		1.992	22:20
Air Blank		0.000	22:21
Control Test		0.081	22:21
Air Blank		0.000	22:21
Diagnostics Check		OK	22:22

Data Download Date/Time: 04/22/2019 10:04

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAFFERTY,CHRISTOPHER,C
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: RAFFERTY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: SPEASE

DATE/TIME OF TEST: 04/08/2019 22:39:36

Diagnostics Check		OK	22:33
Air Blank		0.000	22:34
Control Test		0.081	22:34
Air Blank		0.000	22:35
Subject Sample	# 1	0.000	22:35
Breath Volume		2.343	22:35
Air Blank		0.000	22:35
Air Blank		0.000	22:37
Subject Sample	# 2	0.000	22:38
Breath Volume		2.386	22:38
Air Blank		0.000	22:38
Control Test		0.081	22:39
Air Blank		0.000	22:39
Diagnostics Check		OK	22:39

Data Download Date/Time: 04/22/2019 10:04

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLIDORI,CASEY,A
BREATH TEST OPERATOR AGENCY: Belleview PD
ARRESTING OFFICER: POLIDORI,CASEY
ARRESTING OFFICER AGENCY: Belleview PD
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 437738
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: EBINGER

DATE/TIME OF TEST: 04/09/2019 00:18:16

Diagnostics Check		OK	00:11
Air Blank		0.000	00:12
Control Test		0.081	00:12
Air Blank		0.000	00:13
Subject Sample	# 1	0.182	00:13
Breath Volume		3.621	00:13
Air Blank		0.000	00:14
Air Blank		0.000	00:16
Subject Sample	# 2	0.163	00:16
Breath Volume		1.191	00:16
Air Blank		0.000	00:17
Control Test		0.080	00:17
Air Blank		0.000	00:18
Diagnostics Check		OK	00:18

Data Download Date/Time: 04/22/2019 10:04

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRINGTON,ALEX,M
BREATH TEST OPERATOR AGENCY: Belleview PD
ARRESTING OFFICER: HARRINGTON,ALEX
ARRESTING OFFICER AGENCY: Belleview PD
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: PASTORE

DATE/TIME OF TEST: 04/11/2019 03:32:10

Diagnostics Check		OK	03:20	
Air Blank		0.000	03:21	
Control Test		0.081	03:21	
Air Blank		0.000	03:22	
Subject Sample	# 1		03:25	No Sample Provided
Breath Volume		0.000	03:25	
Air Blank		0.000	03:25	
Air Blank		0.000	03:27	
Subject Sample	# 2		03:30	No Sample Provided
Breath Volume		0.000	03:30	
Air Blank		0.000	03:31	
Control Test		0.081	03:31	
Air Blank		0.000	03:32	
Diagnostics Check		OK	03:32	

Data Download Date/Time: 04/22/2019 10:04

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRINGTON,ALEX,M
BREATH TEST OPERATOR AGENCY: Belleview PD
ARRESTING OFFICER: HARRINGTON,ALEX
ARRESTING OFFICER AGENCY: Belleview PD
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 437738
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: PASTORE

DATE/TIME OF TEST: 04/11/2019 03:37:55

Diagnostics Check		OK	03:33	
Air Blank		0.000	03:34	
Control Test		0.081	03:34	
Air Blank		0.000	03:35	
Subject Sample	# 1		03:36	Subject Test Refused
Breath Volume		0.000	03:36	
Air Blank		0.000	03:37	
Control Test		0.081	03:37	
Air Blank		0.000	03:37	
Diagnostics Check		OK	03:37	

Data Download Date/Time: 04/22/2019 10:04

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ERWIN,R,
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: HAHN,P
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 437738
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:44

SUBJECT'S LAST NAME: MATTHEWS

DATE/TIME OF TEST: 04/12/2019 21:14:03

Diagnostics Check		OK	21:10	
Air Blank		0.000	21:10	
Control Test		0.082	21:10	
Air Blank		0.000	21:11	
Subject Sample	# 1		21:12	Subject Test Refused
Breath Volume		0.000	21:12	
Air Blank		0.000	21:13	
Control Test		0.082	21:13	
Air Blank		0.000	21:13	
Diagnostics Check		OK	21:14	

Data Download Date/Time: 04/22/2019 10:04

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,ALYSSA,D
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: WOJTAS,CHRIS
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 12:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:52

SUBJECT'S LAST NAME: MORAND

DATE/TIME OF TEST: 04/14/2019 13:35:27

Diagnostics Check		OK	13:22	
Air Blank		0.000	13:22	
Control Test		0.080	13:22	
Air Blank		0.000	13:23	
Subject Sample	# 1	0.150	13:26	Volume Not Met
Breath Volume		0.179	13:26	
Air Blank		0.000	13:27	
Air Blank		0.000	13:28	
Subject Sample	# 2	0.187	13:30	
Breath Volume		1.683	13:30	
Air Blank		0.000	13:31	
Air Blank		0.000	13:33	
Subject Sample	# 3	0.193	13:33	
Breath Volume		1.699	13:33	
Air Blank		0.000	13:34	
Control Test		0.080	13:34	
Air Blank		0.000	13:35	
Diagnostics Check		OK	13:35	

Data Download Date/Time: 04/22/2019 10:04

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUELL,KENNETH,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SCHMIDT,ANDREW
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 437738
DRY GAS EXPIRY DATE: 06/08/2019
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: DUELL

DATE/TIME OF TEST: 04/15/2019 01:08:55

Diagnostics Check		OK	01:03	
Air Blank		0.000	01:04	
Control Test		0.081	01:04	
Air Blank		0.000	01:05	
Subject Sample	# 1		01:07	Subject Test Refused
Breath Volume		0.000	01:07	
Air Blank		0.000	01:08	
Control Test		0.081	01:08	
Air Blank		0.000	01:08	
Diagnostics Check		OK	01:08	

Data Download Date/Time: 04/22/2019 10:05

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELRUSSO,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: DELRUSSO,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: GEORGE

DATE/TIME OF TEST: 04/15/2019 21:58:56

Diagnostics Check		OK	21:50
Air Blank		0.000	21:50
Control Test		0.080	21:51
Air Blank		0.000	21:51
Subject Sample	# 1	0.232	21:54
Breath Volume		2.347	21:54
Air Blank		0.000	21:55
Air Blank		0.000	21:56
Subject Sample	# 2	0.218	21:57
Breath Volume		1.917	21:57
Air Blank		0.000	21:58
Control Test		0.080	21:58
Air Blank		0.000	21:58
Diagnostics Check		OK	21:58

Data Download Date/Time: 04/22/2019 10:05

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STREB,KYLE,W
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: OCASIO,FANNIE
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 05:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:10

SUBJECT'S LAST NAME: STREB

DATE/TIME OF TEST: 04/16/2019 06:54:10

Diagnostics Check		OK	06:47
Air Blank		0.000	06:47
Control Test		0.082	06:48
Air Blank		0.000	06:48
Subject Sample	# 1	0.272	06:49
Breath Volume		1.285	06:49
Air Blank		0.000	06:49
Air Blank		0.000	06:51
Subject Sample	# 2	0.286	06:52
Breath Volume		2.355	06:52
Air Blank		0.000	06:53
Control Test		0.081	06:53
Air Blank		0.000	06:54
Diagnostics Check		OK	06:54

Data Download Date/Time: 04/22/2019 10:05

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MERCER,ISAAC,V
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MERCER,ISAAC
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 04/18/2019 02:47:03

Diagnostics Check		OK	02:40
Air Blank		0.000	02:40
Control Test		0.081	02:40
Air Blank		0.000	02:41
Subject Sample	# 1	0.193	02:42
Breath Volume		1.351	02:42
Air Blank		0.000	02:43
Air Blank		0.000	02:44
Subject Sample	# 2	0.187	02:45
Breath Volume		1.320	02:45
Air Blank		0.000	02:46
Control Test		0.081	02:46
Air Blank		0.000	02:46
Diagnostics Check		OK	02:47

Data Download Date/Time: 04/22/2019 10:05

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AUSTIN,NICHOLAS,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: AUSTIN,NICHOLAS
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: QUINN

DATE/TIME OF TEST: 04/18/2019 23:42:54

Diagnostics Check		OK	23:40	
Air Blank		0.000	23:40	
Control Test		0.081	23:40	
Air Blank		0.000	23:41	
Subject Sample	# 1		23:41	Subject Test Refused
Breath Volume		0.000	23:41	
Air Blank		0.000	23:42	
Control Test		0.081	23:42	
Air Blank		0.000	23:42	
Diagnostics Check		OK	23:42	

Data Download Date/Time: 04/22/2019 10:05

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYS,MATTHEW,W
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: POWERS,CHRIS
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 14:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:09

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 04/19/2019 15:36:36

Diagnostics Check		OK	15:29
Air Blank		0.000	15:30
Control Test		0.081	15:30
Air Blank		0.000	15:31
Subject Sample	# 1	0.299	15:31
Breath Volume		2.292	15:31
Air Blank		0.000	15:32
Air Blank		0.000	15:34
Subject Sample	# 2	0.288	15:35
Breath Volume		2.433	15:35
Air Blank		0.000	15:35
Control Test		0.081	15:36
Air Blank		0.000	15:36
Diagnostics Check		OK	15:36

Data Download Date/Time: 04/22/2019 10:05

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO,IVAN,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: SCHMIDT,ANDREW
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/11/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 34917080A3
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: WITKUM

DATE/TIME OF TEST: 04/01/2019 00:40:03

Diagnostics Check		OK	00:33
Air Blank		0.000	00:33
Control Test		0.082	00:34
Air Blank		0.000	00:34
Subject Sample	# 1	0.148	00:35
Breath Volume		1.792	00:35
Air Blank		0.000	00:36
Air Blank		0.000	00:38
Subject Sample	# 2	0.142	00:38
Breath Volume		1.261	00:38
Air Blank		0.000	00:39
Control Test		0.079	00:39
Air Blank		0.000	00:39
Diagnostics Check		OK	00:40

Data Download Date/Time: 04/14/2019 19:21

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,PATRICK,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CASTIGLIONE,T
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A3
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 13:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:41

SUBJECT'S LAST NAME: KRUEGER-BARON

DATE/TIME OF TEST: 04/05/2019 15:04:45

Diagnostics Check		OK	15:01	
Air Blank		0.000	15:01	
Control Test		0.080	15:02	
Air Blank		0.000	15:02	
Subject Sample	# 1		15:03	Subject Test Refused
Breath Volume		0.000	15:03	
Air Blank		0.000	15:03	
Control Test		0.081	15:04	
Air Blank		0.000	15:04	
Diagnostics Check		OK	15:04	

Data Download Date/Time: 04/14/2019 19:21

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROFT,BRETT,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CROFT,BRETT
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 03/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A3
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: CHEEVER

DATE/TIME OF TEST: 04/08/2019 22:18:16

Diagnostics Check		OK	22:13	
Air Blank		0.000	22:14	
Control Test		0.081	22:14	
Air Blank		0.000	22:15	
Subject Sample	# 1		22:16	Subject Test Refused
Breath Volume		0.000	22:16	
Air Blank		0.000	22:17	
Control Test		0.080	22:17	
Air Blank		0.000	22:18	
Diagnostics Check		OK	22:18	

Data Download Date/Time: 04/14/2019 19:21

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROFT,BRETT,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CROFT,BRETT
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 03/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A3
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: PFLEGER

DATE/TIME OF TEST: 04/10/2019 00:58:51

Diagnostics Check		OK	00:51
Air Blank		0.000	00:51
Control Test		0.082	00:52
Air Blank		0.000	00:52
Subject Sample	# 1	0.076	00:54
Breath Volume		2.867	00:54
Air Blank		0.000	00:54
Air Blank		0.000	00:56
Subject Sample	# 2	0.079	00:57
Breath Volume		2.859	00:57
Air Blank		0.000	00:57
Control Test		0.081	00:58
Air Blank		0.000	00:58
Diagnostics Check		OK	00:58

Data Download Date/Time: 04/14/2019 19:21

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROFT,BRETT,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CROFT,BRETT
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 03/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A3
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 04/14/2019 02:28:45

Diagnostics Check		OK	02:21
Air Blank		0.000	02:22
Control Test		0.082	02:22
Air Blank		0.000	02:23
Subject Sample	# 1	0.000	02:24
Breath Volume		3.332	02:24
Air Blank		0.000	02:24
Air Blank		0.000	02:26
Subject Sample	# 2	0.000	02:27
Breath Volume		2.917	02:27
Air Blank		0.000	02:27
Control Test		0.081	02:28
Air Blank		0.000	02:28
Diagnostics Check		OK	02:28

Data Download Date/Time: 04/14/2019 19:21

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROFT,BRETT,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CROFT,BRETT
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A3
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: KAIROS

DATE/TIME OF TEST: 04/15/2019 00:58:59

Diagnostics Check		OK	00:56	
Air Blank		0.000	00:56	
Control Test		0.081	00:56	
Air Blank		0.000	00:57	
Subject Sample	# 1		00:57	Subject Test Refused
Breath Volume		0.000	00:57	
Air Blank		0.000	00:58	
Control Test		0.081	00:58	
Air Blank		0.000	00:58	
Diagnostics Check		OK	00:58	

Data Download Date/Time: 05/09/2019 13:19

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROFT,BRETT,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CROFT,BRETT
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A3
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: BEAN

DATE/TIME OF TEST: 04/18/2019 01:15:42

Diagnostics Check		OK	01:12	
Air Blank		0.000	01:12	
Control Test		0.081	01:13	
Air Blank		0.000	01:13	
Subject Sample	# 1		01:14	Subject Test Refused
Breath Volume		0.000	01:14	
Air Blank		0.000	01:14	
Control Test		0.081	01:15	
Air Blank		0.000	01:15	
Diagnostics Check		OK	01:15	

Data Download Date/Time: 05/09/2019 13:19

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEADOWS,GEOFFREY,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: KIBBEY,BRADLEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A3
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:36

SUBJECT'S LAST NAME: BARKLEY

DATE/TIME OF TEST: 04/19/2019 21:08:53

Diagnostics Check		OK	20:56	
Air Blank		0.000	20:56	
Control Test		0.082	20:57	
Air Blank		0.000	20:57	
Subject Sample	# 1	0.177	20:58	
Breath Volume		1.617	20:58	
Air Blank		0.000	20:59	
Air Blank		0.000	21:00	
Subject Sample	# 2		21:03	No Sample Provided
Breath Volume		0.000	21:03	
Air Blank		0.000	21:04	
Air Blank		0.000	21:06	
Subject Sample	# 3	0.180	21:07	
Breath Volume		1.699	21:07	
Air Blank		0.000	21:07	
Control Test		0.082	21:08	
Air Blank		0.000	21:08	
Diagnostics Check		OK	21:08	

Data Download Date/Time: 05/09/2019 13:19

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO,IVAN,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MITRO,JOSHUA
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/14/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 34917080A3
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:41

SUBJECT'S LAST NAME: BIDLACK

DATE/TIME OF TEST: 04/22/2019 22:08:10

Diagnostics Check		OK	22:01
Air Blank		0.000	22:02
Control Test		0.082	22:02
Air Blank		0.000	22:02
Subject Sample	# 1	0.106	22:03
Breath Volume		4.183	22:03
Air Blank		0.000	22:04
Air Blank		0.000	22:06
Subject Sample	# 2	0.097	22:06
Breath Volume		2.242	22:06
Air Blank		0.000	22:07
Control Test		0.081	22:07
Air Blank		0.000	22:08
Diagnostics Check		OK	22:08

Data Download Date/Time: 05/09/2019 13:19

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROFT,BRETT,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CROFT,BRETT
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A3
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 04/24/2019 00:03:04

Diagnostics Check		OK	23:54
Air Blank		0.000	23:54
Control Test		0.081	23:54
Air Blank		0.000	23:55
Subject Sample	# 2	0.163	23:57
Breath Volume		2.171	23:57
Air Blank		0.000	23:58
Air Blank		0.000	00:00
Subject Sample	# 1	0.162	00:01
Breath Volume		2.683	00:01
Air Blank		0.000	00:02
Control Test		0.080	00:02
Air Blank		0.000	00:02
Diagnostics Check		OK	00:03

Data Download Date/Time: 05/09/2019 13:19

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KJEKSTAD,ERIC,C
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KJEKSTAD,ERIC
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A3
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: HERNANDEZ VEGA

DATE/TIME OF TEST: 04/24/2019 03:53:09

Diagnostics Check		OK	03:45
Air Blank		0.000	03:45
Control Test		0.081	03:46
Air Blank		0.000	03:46
Subject Sample	# 1	0.099	03:48
Breath Volume		2.914	03:48
Air Blank		0.000	03:49
Air Blank		0.000	03:51
Subject Sample	# 2	0.098	03:51
Breath Volume		2.984	03:51
Air Blank		0.000	03:52
Control Test		0.079	03:52
Air Blank		0.000	03:53
Diagnostics Check		OK	03:53

Data Download Date/Time: 05/09/2019 13:19

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,PATRICK,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: GROVES,D
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A3
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 05:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:58

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 04/25/2019 06:22:42

Diagnostics Check		OK	06:18	
Air Blank		0.000	06:19	
Control Test		0.081	06:19	
Air Blank		0.000	06:19	
Subject Sample	# 1		06:21	Subject Test Refused
Breath Volume		0.000	06:21	
Air Blank		0.000	06:21	
Control Test		0.081	06:22	
Air Blank		0.000	06:22	
Diagnostics Check		OK	06:22	

Data Download Date/Time: 05/09/2019 13:19

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROFT,BRETT,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CROFT,BRETT
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A3
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: DECKER

DATE/TIME OF TEST: 04/27/2019 03:22:02

Diagnostics Check		OK	03:17	
Air Blank		0.000	03:18	
Control Test		0.081	03:18	
Air Blank		0.000	03:18	
Subject Sample	# 1		03:20	Subject Test Refused
Breath Volume		0.000	03:20	
Air Blank		0.000	03:21	
Control Test		0.081	03:21	
Air Blank		0.000	03:21	
Diagnostics Check		OK	03:22	

Data Download Date/Time: 05/09/2019 13:19

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTIZ,B,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ORTIZ,B
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A3
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: NEW

DATE/TIME OF TEST: 04/28/2019 01:22:14

Diagnostics Check		OK	01:15
Air Blank		0.000	01:16
Control Test		0.081	01:16
Air Blank		0.000	01:17
Subject Sample	# 1	0.103	01:17
Breath Volume		1.980	01:17
Air Blank		0.000	01:18
Air Blank		0.000	01:20
Subject Sample	# 2	0.103	01:20
Breath Volume		1.628	01:20
Air Blank		0.000	01:21
Control Test		0.081	01:21
Air Blank		0.000	01:22
Diagnostics Check		OK	01:22

Data Download Date/Time: 05/09/2019 13:19

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLLITO,BRAD,R
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MITRO,JOSHUA
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A3
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 04/29/2019 01:51:31

Diagnostics Check		OK	01:48	
Air Blank		0.000	01:48	
Control Test		0.082	01:49	
Air Blank		0.000	01:49	
Subject Sample	# 1		01:50	Subject Test Refused
Breath Volume		0.000	01:50	
Air Blank		0.000	01:50	
Control Test		0.081	01:50	
Air Blank		0.000	01:51	
Diagnostics Check		OK	01:51	

Data Download Date/Time: 05/09/2019 13:19

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUTTON,EDWIN,C
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: TAN,R
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 04/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:27

SUBJECT'S LAST NAME: FROST

DATE/TIME OF TEST: 04/28/2019 18:56:09

Diagnostics Check		OK	18:49
Air Blank		0.000	18:49
Control Test		0.079	18:50
Air Blank		0.000	18:50
Subject Sample	# 1	0.203	18:51
Breath Volume		1.187	18:51
Air Blank		0.000	18:51
Air Blank		0.000	18:53
Subject Sample	# 2	0.216	18:54
Breath Volume		3.367	18:54
Air Blank		0.000	18:55
Control Test		0.077	18:55
Air Blank		0.000	18:56
Diagnostics Check		OK	18:56

Data Download Date/Time: 04/30/2019 17:07

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,JARROD,W
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: SMITH,CHARLES
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 04/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: HOLLAR

DATE/TIME OF TEST: 04/30/2019 03:23:26

Diagnostics Check		OK	03:20	
Air Blank		0.000	03:21	
Control Test		0.079	03:21	
Air Blank		0.000	03:21	
Subject Sample	# 1		03:22	Subject Test Refused
Breath Volume		0.000	03:22	
Air Blank		0.000	03:22	
Control Test		0.079	03:22	
Air Blank		0.000	03:23	
Diagnostics Check		OK	03:23	

Data Download Date/Time: 04/30/2019 17:07

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LETSON,REBECCA,G
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: LETSON,REBECCA
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 04/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 06:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:15

SUBJECT'S LAST NAME: COBB

DATE/TIME OF TEST: 04/30/2019 07:43:22

Diagnostics Check		OK	07:36
Air Blank		0.000	07:37
Control Test		0.079	07:37
Air Blank		0.000	07:37
Subject Sample	# 1	0.000	07:38
Breath Volume		2.746	07:38
Air Blank		0.000	07:39
Air Blank		0.000	07:41
Subject Sample	# 2	0.000	07:41
Breath Volume		2.789	07:41
Air Blank		0.000	07:42
Control Test		0.078	07:42
Air Blank		0.000	07:43
Diagnostics Check		OK	07:43

Data Download Date/Time: 04/30/2019 17:07

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLIS,TERENCE,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: BERKIKISER,R
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 03/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 06:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:21

SUBJECT'S LAST NAME: STAFFORD

DATE/TIME OF TEST: 04/01/2019 07:44:31

Diagnostics Check		OK	07:41	
Air Blank		0.000	07:42	
Control Test		0.079	07:42	
Air Blank		0.000	07:42	
Subject Sample	# 1		07:43	Subject Test Refused
Breath Volume		0.000	07:43	
Air Blank		0.000	07:43	
Control Test		0.079	07:43	
Air Blank		0.000	07:44	
Diagnostics Check		OK	07:44	

Data Download Date/Time: 04/30/2019 17:05

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLIS,TERENCE,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: BERKIKISER,R
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 03/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 06:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:29

SUBJECT'S LAST NAME: STAFFORD

DATE/TIME OF TEST: 04/01/2019 09:56:06

Diagnostics Check		OK	09:49
Air Blank		0.000	09:50
Control Test		0.079	09:50
Air Blank		0.000	09:51
Subject Sample	# 1	0.000	09:51
Breath Volume		4.054	09:51
Air Blank		0.000	09:52
Air Blank		0.000	09:54
Subject Sample	# 2	0.000	09:54
Breath Volume		3.625	09:54
Air Blank		0.000	09:55
Control Test		0.080	09:55
Air Blank		0.000	09:56
Diagnostics Check		OK	09:56

Data Download Date/Time: 04/30/2019 17:05

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BISHOP METZ,FELICIA,B
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: SLAY,ADRIAN
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 03/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: BARNWELL

DATE/TIME OF TEST: 04/05/2019 00:11:02

Diagnostics Check		OK	23:59
Air Blank		0.000	00:00
Control Test		0.079	00:00
Air Blank		0.000	00:01
Subject Sample	# 1	0.167	00:03
Breath Volume		1.546	00:03
Air Blank		0.000	00:04
Air Blank		0.000	00:05
Subject Sample	# 2	0.178	00:09
Breath Volume		2.660	00:09
Air Blank		0.000	00:10
Control Test		0.078	00:10
Air Blank		0.000	00:10
Diagnostics Check		OK	00:11

Data Download Date/Time: 04/30/2019 17:05

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARSHALL,JARROD,C
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: RAFFERTY,C
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 03/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:48

SUBJECT'S LAST NAME: MEYER

DATE/TIME OF TEST: 04/06/2019 19:15:02

Diagnostics Check		OK	19:08
Air Blank		0.000	19:08
Control Test		0.079	19:09
Air Blank		0.000	19:09
Subject Sample	# 1	0.000	19:10
Breath Volume		1.191	19:10
Air Blank		0.000	19:11
Air Blank		0.000	19:13
Subject Sample	# 2	0.000	19:13
Breath Volume		1.152	19:13
Air Blank		0.000	19:14
Control Test		0.079	19:14
Air Blank		0.000	19:14
Diagnostics Check		OK	19:15

Data Download Date/Time: 04/30/2019 17:05

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLIS,TERENCE,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: HOLLOWAY,J
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 03/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:36

SUBJECT'S LAST NAME: PISTILLI

DATE/TIME OF TEST: 04/07/2019 17:03:10

Diagnostics Check		OK	16:56
Air Blank		0.000	16:57
Control Test		0.079	16:57
Air Blank		0.000	16:57
Subject Sample	# 1	0.000	16:58
Breath Volume		2.429	16:58
Air Blank		0.000	16:59
Air Blank		0.000	17:01
Subject Sample	# 2	0.000	17:01
Breath Volume		2.265	17:01
Air Blank		0.000	17:02
Control Test		0.079	17:02
Air Blank		0.000	17:03
Diagnostics Check		OK	17:03

Data Download Date/Time: 04/30/2019 17:05

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BISHOP-METZ,FELICIA,B
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: SMITH,CHARLES
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 03/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 04/08/2019 20:04:27

Diagnostics Check		OK	20:01	
Air Blank		0.000	20:01	
Control Test		0.078	20:02	
Air Blank		0.000	20:02	
Subject Sample	# 1		20:03	Subject Test Refused
Breath Volume		0.000	20:03	
Air Blank		0.000	20:03	
Control Test		0.078	20:03	
Air Blank		0.000	20:04	
Diagnostics Check		OK	20:04	

Data Download Date/Time: 04/30/2019 17:05

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BISHOP-METZ,FELICIA,B
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: HINDERHOFER,EDWARD
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 03/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: DEMAREST

DATE/TIME OF TEST: 04/09/2019 01:28:00

Diagnostics Check		OK	01:17	
Air Blank		0.000	01:18	
Control Test		0.078	01:18	
Air Blank		0.000	01:19	
Subject Sample	# 1	0.141	01:22	Volume Not Met
Breath Volume		0.734	01:22	
Air Blank		0.000	01:22	
Air Blank		0.000	01:24	
Subject Sample	# 2		01:26	Subject Test Refused
Breath Volume		0.039	01:26	
Air Blank		0.000	01:27	
Control Test		0.077	01:27	
Air Blank		0.000	01:27	
Diagnostics Check		OK	01:28	

Data Download Date/Time: 04/30/2019 17:05

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LETSON,REBECCA,G
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: LETSON,REBECCA
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 03/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 12:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:09

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 04/11/2019 13:36:19

Diagnostics Check		OK	13:30
Air Blank		0.000	13:30
Control Test		0.079	13:30
Air Blank		0.000	13:31
Subject Sample	# 1	0.150	13:31
Breath Volume		2.925	13:31
Air Blank		0.000	13:32
Air Blank		0.000	13:34
Subject Sample	# 2	0.147	13:34
Breath Volume		2.617	13:34
Air Blank		0.000	13:35
Control Test		0.077	13:35
Air Blank		0.000	13:36
Diagnostics Check		OK	13:36

Data Download Date/Time: 04/30/2019 17:05

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUTTON,EDWIN,C
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: SLAY,A
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 03/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 04/13/2019 23:49:58

Diagnostics Check		OK	23:40
Air Blank		0.000	23:41
Control Test		0.078	23:41
Air Blank		0.000	23:42
Subject Sample	# 1	0.155	23:45
Breath Volume		3.257	23:45
Air Blank		0.000	23:45
Air Blank		0.000	23:47
Subject Sample	# 2	0.142	23:48
Breath Volume		1.996	23:48
Air Blank		0.000	23:49
Control Test		0.077	23:49
Air Blank		0.000	23:49
Diagnostics Check		OK	23:49

Data Download Date/Time: 04/30/2019 17:05

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUTTON,EDWIN,C
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: SMITH,C
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 03/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: BEASLEY

DATE/TIME OF TEST: 04/14/2019 00:17:24

Diagnostics Check		OK	00:09
Air Blank		0.000	00:10
Control Test		0.078	00:10
Air Blank		0.000	00:11
Subject Sample	# 1	0.140	00:12
Breath Volume		2.789	00:12
Air Blank		0.000	00:13
Air Blank		0.000	00:15
Subject Sample	# 2	0.141	00:15
Breath Volume		3.066	00:15
Air Blank		0.000	00:16
Control Test		0.077	00:16
Air Blank		0.000	00:17
Diagnostics Check		OK	00:17

Data Download Date/Time: 04/30/2019 17:05

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BISHOP-METZ,FELICIA,B
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: SCHMIDT,ANDREW
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: HAFER

DATE/TIME OF TEST: 04/17/2019 23:58:15

Diagnostics Check		OK	23:55	
Air Blank		0.000	23:55	
Control Test		0.079	23:56	
Air Blank		0.000	23:56	
Subject Sample	# 1		23:56	Subject Test Refused
Breath Volume		0.000	23:56	
Air Blank		0.000	23:57	
Control Test		0.079	23:57	
Air Blank		0.000	23:58	
Diagnostics Check		OK	23:58	

Data Download Date/Time: 04/30/2019 17:05

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLIS,TERENCE,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: RAFFERTY,C
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 03/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 14:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:39

SUBJECT'S LAST NAME: MCDONALD

DATE/TIME OF TEST: 04/20/2019 16:08:42

Diagnostics Check		OK	16:02
Air Blank		0.000	16:02
Control Test		0.079	16:03
Air Blank		0.000	16:03
Subject Sample	# 1	0.000	16:04
Breath Volume		3.050	16:04
Air Blank		0.000	16:04
Air Blank		0.000	16:06
Subject Sample	# 2	0.000	16:07
Breath Volume		2.777	16:07
Air Blank		0.000	16:07
Control Test		0.079	16:08
Air Blank		0.000	16:08
Diagnostics Check		OK	16:08

Data Download Date/Time: 04/30/2019 17:05

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,JARROD,W
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: LOMBARDI,ANTHONY
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 03/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: PAILET

DATE/TIME OF TEST: 04/24/2019 22:10:32

Diagnostics Check		OK	22:04
Air Blank		0.000	22:04
Control Test		0.078	22:04
Air Blank		0.000	22:05
Subject Sample	# 1	0.202	22:05
Breath Volume		1.843	22:05
Air Blank		0.000	22:06
Air Blank		0.000	22:08
Subject Sample	# 2	0.212	22:08
Breath Volume		1.996	22:08
Air Blank		0.000	22:09
Control Test		0.078	22:09
Air Blank		0.000	22:10
Diagnostics Check		OK	22:10

Data Download Date/Time: 04/30/2019 17:05

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHOETTNER,ADRIAN,J
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: HOLTZHOUSE,J
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:42

SUBJECT'S LAST NAME: MOLITOR

DATE/TIME OF TEST: 04/03/2019 19:10:46

Diagnostics Check		OK	19:03
Air Blank		0.000	19:04
Control Test		0.077	19:04
Air Blank		0.000	19:04
Subject Sample	# 1	0.291	19:05
Breath Volume		1.828	19:05
Air Blank		0.000	19:06
Air Blank		0.000	19:08
Subject Sample	# 2	0.271	19:09
Breath Volume		1.164	19:09
Air Blank		0.000	19:09
Control Test		0.076	19:10
Air Blank		0.000	19:10
Diagnostics Check		OK	19:10

Data Download Date/Time: 05/01/2019 07:00

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVENPORT,JOSHUA,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DAVENPORT,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: SHAFFER

DATE/TIME OF TEST: 04/04/2019 01:07:15

Diagnostics Check		OK	01:00
Air Blank		0.000	01:01
Control Test		0.077	01:01
Air Blank		0.000	01:02
Subject Sample	# 1	0.208	01:02
Breath Volume		1.585	01:02
Air Blank		0.000	01:03
Air Blank		0.000	01:05
Subject Sample	# 2	0.213	01:05
Breath Volume		1.550	01:05
Air Blank		0.000	01:06
Control Test		0.077	01:06
Air Blank		0.000	01:07
Diagnostics Check		OK	01:07

Data Download Date/Time: 05/01/2019 07:00

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ECHEVARRIA,LISA,V
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: WIGGLESWORTH,C
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: BARKER

DATE/TIME OF TEST: 04/07/2019 04:32:31

Diagnostics Check		OK	04:29	
Air Blank		0.000	04:30	
Control Test		0.077	04:30	
Air Blank		0.000	04:30	
Subject Sample	# 1		04:31	Subject Test Refused
Breath Volume		0.000	04:31	
Air Blank		0.000	04:31	
Control Test		0.079	04:31	
Air Blank		0.000	04:32	
Diagnostics Check		OK	04:32	

Data Download Date/Time: 05/01/2019 07:00

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,STEVEN,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WARD,S
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: PELLEGRINO

DATE/TIME OF TEST: 04/10/2019 23:11:52

Diagnostics Check		OK	23:08	
Air Blank		0.000	23:09	
Control Test		0.078	23:09	
Air Blank		0.000	23:10	
Subject Sample	# 1		23:10	Subject Test Refused
Breath Volume		0.000	23:10	
Air Blank		0.000	23:10	
Control Test		0.078	23:11	
Air Blank		0.000	23:11	
Diagnostics Check		OK	23:11	

Data Download Date/Time: 05/01/2019 07:00

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ECHEVARRIA,LISA,V
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: WIGGLESWORTH,C
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: MUNNERLYN

DATE/TIME OF TEST: 04/11/2019 00:56:17

Diagnostics Check		OK	00:49
Air Blank		0.000	00:49
Control Test		0.078	00:50
Air Blank		0.000	00:50
Subject Sample	# 1	0.000	00:51
Breath Volume		2.804	00:51
Air Blank		0.000	00:52
Air Blank		0.000	00:54
Subject Sample	# 2	0.000	00:54
Breath Volume		2.593	00:54
Air Blank		0.000	00:55
Control Test		0.079	00:55
Air Blank		0.000	00:56
Diagnostics Check		OK	00:56

Data Download Date/Time: 05/01/2019 07:00

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWIGER,CARL,M
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: DOLAND,T
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:25

SUBJECT'S LAST NAME: DELLINGER

DATE/TIME OF TEST: 04/22/2019 17:02:39

Diagnostics Check		OK	16:56
Air Blank		0.000	16:56
Control Test		0.078	16:57
Air Blank		0.000	16:57
Subject Sample	# 1	0.000	16:58
Breath Volume		2.167	16:58
Air Blank		0.000	16:58
Air Blank		0.000	17:00
Subject Sample	# 2	0.000	17:01
Breath Volume		2.437	17:01
Air Blank		0.000	17:01
Control Test		0.078	17:02
Air Blank		0.000	17:02
Diagnostics Check		OK	17:02

Data Download Date/Time: 05/01/2019 07:00

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ECHEVARRIA,LISA,V
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: FERRER,F
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: KANOUSE

DATE/TIME OF TEST: 04/25/2019 00:45:42

Diagnostics Check		OK	00:42	
Air Blank		0.000	00:43	
Control Test		0.077	00:43	
Air Blank		0.000	00:44	
Subject Sample	# 1		00:44	Subject Test Refused
Breath Volume		0.000	00:44	
Air Blank		0.000	00:44	
Control Test		0.078	00:45	
Air Blank		0.000	00:45	
Diagnostics Check		OK	00:45	

Data Download Date/Time: 05/01/2019 07:00

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DADDARIO,A,
BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)
ARRESTING OFFICER: REED,T
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: PELOQUIN

DATE/TIME OF TEST: 04/27/2019 00:59:58

Diagnostics Check		OK	00:57	
Air Blank		0.000	00:57	
Control Test		0.077	00:57	
Air Blank		0.000	00:58	
Subject Sample	# 1		00:58	Subject Test Refused
Breath Volume		0.000	00:58	
Air Blank		0.000	00:59	
Control Test		0.078	00:59	
Air Blank		0.000	00:59	
Diagnostics Check		OK	00:59	

Data Download Date/Time: 05/01/2019 07:00

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DADDARIO,ASHLEY,A
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: SCHMIDT,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: PRECHTL

DATE/TIME OF TEST: 04/27/2019 01:40:56

Diagnostics Check		OK	01:27	
Air Blank		0.000	01:28	
Control Test		0.078	01:28	
Air Blank		0.000	01:28	
Subject Sample	# 1		01:31	No Sample Provided
Breath Volume		0.000	01:31	
Air Blank		0.000	01:32	
Air Blank		0.000	01:34	
Subject Sample	# 2	0.154	01:35	
Breath Volume		1.121	01:35	
Air Blank		0.000	01:36	
Air Blank		0.000	01:37	
Subject Sample	# 3	0.150	01:39	
Breath Volume		1.410	01:39	
Air Blank		0.000	01:40	
Control Test		0.077	01:40	
Air Blank		0.000	01:40	
Diagnostics Check		OK	01:40	

Data Download Date/Time: 05/01/2019 07:00

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KJEKSTAD,ERIC,C
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KJEKSTAD,ERIC
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 04/27/2019 21:19:45

Diagnostics Check		OK	21:13
Air Blank		0.000	21:13
Control Test		0.077	21:14
Air Blank		0.000	21:14
Subject Sample	# 1	0.169	21:15
Breath Volume		1.914	21:15
Air Blank		0.000	21:15
Air Blank		0.000	21:17
Subject Sample	# 2	0.168	21:18
Breath Volume		2.316	21:18
Air Blank		0.000	21:18
Control Test		0.077	21:19
Air Blank		0.000	21:19
Diagnostics Check		OK	21:19

Data Download Date/Time: 05/01/2019 07:00

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINDER,BRONSON,M
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: BINDER,BRONSON
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: GOODWIN

DATE/TIME OF TEST: 04/04/2019 23:54:45

Diagnostics Check		OK	23:48	
Air Blank		0.000	23:48	
Control Test		0.077	23:49	
Air Blank		0.000	23:49	
Subject Sample	# 1		23:52	No Sample Provided
Breath Volume		0.000	23:52	
Air Blank		0.000	23:53	
Air Blank		0.000	23:54	

Data Download Date/Time: 04/10/2019 23:47

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINDER,BRONSON,M
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: BINDER,BRONSON
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: GOODWIN

DATE/TIME OF TEST: 04/05/2019 00:06:45

Diagnostics Check		OK	00:00
Air Blank		0.000	00:00
Control Test		0.076	00:00
Air Blank		0.000	00:01
Subject Sample	# 1	0.185	00:01
Breath Volume		3.996	00:01
Air Blank		0.000	00:02
Air Blank		0.000	00:04
Subject Sample	# 2	0.189	00:05
Breath Volume		3.207	00:05
Air Blank		0.000	00:05
Control Test		0.077	00:06
Air Blank		0.000	00:06
Diagnostics Check		OK	00:06

Data Download Date/Time: 04/10/2019 23:47

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINDER,BRONSON,M
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CHESSHER,SANDRA
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: MOBLEY

DATE/TIME OF TEST: 04/06/2019 03:14:55

Diagnostics Check		OK	03:08
Air Blank		0.000	03:08
Control Test		0.076	03:09
Air Blank		0.000	03:09
Subject Sample	# 1	0.115	03:10
Breath Volume		2.167	03:10
Air Blank		0.000	03:10
Air Blank		0.000	03:12
Subject Sample	# 2	0.120	03:13
Breath Volume		2.097	03:13
Air Blank		0.000	03:14
Control Test		0.077	03:14
Air Blank		0.000	03:14
Diagnostics Check		OK	03:14

Data Download Date/Time: 04/10/2019 23:47

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHESSHER,SANDRA,W
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CROW,S
ARRESTING OFFICER AGENCY: Umatilla PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:26

SUBJECT'S LAST NAME: SUGDEN

DATE/TIME OF TEST: 04/10/2019 21:53:42

Diagnostics Check		OK	21:47
Air Blank		0.000	21:47
Control Test		0.076	21:47
Air Blank		0.000	21:48
Subject Sample	# 1	0.124	21:48
Breath Volume		1.839	21:48
Air Blank		0.000	21:49
Air Blank		0.000	21:51
Subject Sample	# 2	0.121	21:52
Breath Volume		1.785	21:52
Air Blank		0.000	21:52
Control Test		0.077	21:53
Air Blank		0.000	21:53
Diagnostics Check		OK	21:53

Data Download Date/Time: 04/10/2019 23:47

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITCHER,DARRYL,
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: BINDER,BRONSON
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: ARABIA

DATE/TIME OF TEST: 04/11/2019 02:47:02

Diagnostics Check		OK	02:40
Air Blank		0.000	02:40
Control Test		0.077	02:41
Air Blank		0.000	02:41
Subject Sample	# 1	0.180	02:42
Breath Volume		3.007	02:42
Air Blank		0.000	02:42
Air Blank		0.000	02:44
Subject Sample	# 2	0.185	02:45
Breath Volume		3.078	02:45
Air Blank		0.000	02:46
Control Test		0.078	02:46
Air Blank		0.000	02:46
Diagnostics Check		OK	02:47

Data Download Date/Time: 05/16/2019 16:40

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITCHER,DARRYL,P
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: BINDER,BRONSON
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 04/12/2019 02:22:16

Diagnostics Check		OK	02:15
Air Blank		0.000	02:16
Control Test		0.078	02:16
Air Blank		0.000	02:17
Subject Sample	# 1	0.194	02:17
Breath Volume		1.984	02:17
Air Blank		0.000	02:18
Air Blank		0.000	02:20
Subject Sample	# 2	0.196	02:20
Breath Volume		1.917	02:20
Air Blank		0.000	02:21
Control Test		0.077	02:21
Air Blank		0.000	02:22
Diagnostics Check		OK	02:22

Data Download Date/Time: 05/16/2019 16:40

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITCHER,DARRYL,P
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: BINDER,BRONSON
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: BARRINER

DATE/TIME OF TEST: 04/13/2019 02:55:31

Diagnostics Check		OK	02:47
Air Blank		0.000	02:48
Control Test		0.077	02:48
Air Blank		0.000	02:49
Subject Sample	# 1	0.133	02:50
Breath Volume		2.605	02:50
Air Blank		0.000	02:51
Air Blank		0.000	02:53
Subject Sample	# 2	0.136	02:53
Breath Volume		1.820	02:53
Air Blank		0.000	02:54
Control Test		0.078	02:54
Air Blank		0.000	02:55
Diagnostics Check		OK	02:55

Data Download Date/Time: 05/16/2019 16:40

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINDER,BRONSON,M
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: PITCHER,DARRYL
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 04/14/2019 01:22:31

Diagnostics Check		OK	01:15
Air Blank		0.000	01:16
Control Test		0.077	01:16
Air Blank		0.000	01:17
Subject Sample	# 1	0.133	01:17
Breath Volume		1.789	01:17
Air Blank		0.000	01:18
Air Blank		0.000	01:20
Subject Sample	# 2	0.126	01:20
Breath Volume		1.789	01:20
Air Blank		0.000	01:21
Control Test		0.077	01:21
Air Blank		0.000	01:22
Diagnostics Check		OK	01:22

Data Download Date/Time: 05/16/2019 16:40

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINDER,BRONSON,M
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: BINDER,BRONSON
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: KAUFFMAN

DATE/TIME OF TEST: 04/14/2019 02:23:38

Diagnostics Check		OK	02:12	
Air Blank		0.000	02:12	
Control Test		0.076	02:12	
Air Blank		0.000	02:13	
Subject Sample	# 1		02:16	No Sample Provided
Breath Volume		0.000	02:16	
Air Blank		0.000	02:17	
Air Blank		0.000	02:19	
Subject Sample	# 2		02:22	No Sample Provided
Breath Volume		0.000	02:22	
Air Blank		0.000	02:22	
Control Test		0.076	02:23	
Air Blank		0.000	02:23	
Diagnostics Check		OK	02:23	

Data Download Date/Time: 05/16/2019 16:40

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINDER,BRONSON,M
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CHESSHER,SANDRA
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: MCRAY

DATE/TIME OF TEST: 04/20/2019 00:40:50

Diagnostics Check		OK	00:31
Air Blank		0.000	00:32
Control Test		0.078	00:32
Air Blank		0.000	00:33
Subject Sample	# 1	0.124	00:35
Breath Volume		1.222	00:35
Air Blank		0.000	00:36
Air Blank		0.000	00:38
Subject Sample	# 2	0.120	00:39
Breath Volume		1.207	00:39
Air Blank		0.000	00:39
Control Test		0.076	00:40
Air Blank		0.000	00:40
Diagnostics Check		OK	00:40

Data Download Date/Time: 05/16/2019 16:40

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINDER,BRONSON,M
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: PITCHER,DARRYL
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: KIES

DATE/TIME OF TEST: 04/20/2019 01:46:39

Diagnostics Check		OK	01:39
Air Blank		0.000	01:40
Control Test		0.077	01:40
Air Blank		0.000	01:40
Subject Sample	# 1	0.192	01:41
Breath Volume		2.687	01:41
Air Blank		0.000	01:42
Air Blank		0.000	01:44
Subject Sample	# 2	0.188	01:45
Breath Volume		2.460	01:45
Air Blank		0.000	01:45
Control Test		0.075	01:46
Air Blank		0.000	01:46
Diagnostics Check		OK	01:46

Data Download Date/Time: 05/16/2019 16:40

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHESSHER,SANDRA,W
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: BINDER,B
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: ROBBINS

DATE/TIME OF TEST: 04/20/2019 03:33:44

Diagnostics Check		OK	03:18	
Air Blank		0.000	03:18	
Control Test		0.077	03:18	
Air Blank		0.000	03:19	
Subject Sample	# 1	0.112	03:22	Volume Not Met
Breath Volume		0.761	03:22	
Air Blank		0.000	03:23	
Air Blank		0.000	03:25	
Subject Sample	# 2	0.130	03:27	
Breath Volume		1.664	03:27	
Air Blank		0.000	03:27	
Air Blank		0.000	03:29	
Subject Sample	# 3	0.118	03:32	
Breath Volume		3.152	03:32	
Air Blank		0.000	03:32	
Control Test		0.075	03:33	
Air Blank		0.000	03:33	
Diagnostics Check		OK	03:33	

Data Download Date/Time: 05/16/2019 16:40

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINDER,BRONSON,M
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: BINDER,BRONSON
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: RECORD

DATE/TIME OF TEST: 04/26/2019 03:11:20

Diagnostics Check		OK	03:02	
Air Blank		0.000	03:03	
Control Test		0.076	03:03	
Air Blank		0.000	03:04	
Subject Sample	# 1		03:07	No Sample Provided
Breath Volume		0.000	03:07	
Air Blank		0.000	03:07	
Air Blank		0.000	03:09	
Subject Sample	# 2		03:09	Subject Test Refused
Breath Volume		0.000	03:09	
Air Blank		0.000	03:10	
Control Test		0.077	03:10	
Air Blank		0.000	03:11	
Diagnostics Check		OK	03:11	

Data Download Date/Time: 05/16/2019 16:40

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINDER,BRONSON,M
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: BINDER,BRONSON
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 22:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: CAUDILL

DATE/TIME OF TEST: 04/26/2019 22:34:56

Diagnostics Check		OK	22:24
Air Blank		0.000	22:24
Control Test		0.076	22:25
Air Blank		0.000	22:25
Subject Sample	# 1	0.186	22:28
Breath Volume		1.253	22:28
Air Blank		0.000	22:29
Air Blank		0.000	22:31
Subject Sample	# 2	0.193	22:33
Breath Volume		1.136	22:33
Air Blank		0.000	22:34
Control Test		0.077	22:34
Air Blank		0.000	22:34
Diagnostics Check		OK	22:34

Data Download Date/Time: 05/16/2019 16:40

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRUSE,KRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: CHESSHER,SANDI
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 04/27/2019 01:54:47

Diagnostics Check		OK	01:47
Air Blank		0.000	01:48
Control Test		0.078	01:48
Air Blank		0.000	01:49
Subject Sample	# 1	0.115	01:49
Breath Volume		2.449	01:49
Air Blank		0.000	01:50
Air Blank		0.000	01:52
Subject Sample	# 2	0.106	01:53
Breath Volume		2.351	01:53
Air Blank		0.000	01:53
Control Test		0.078	01:54
Air Blank		0.000	01:54
Diagnostics Check		OK	01:54

Data Download Date/Time: 05/16/2019 16:40

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINDER,BRONSON,M
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: RAFFERTY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 19:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: VALERO RODRIGUEZ

DATE/TIME OF TEST: 04/27/2019 20:37:15

Diagnostics Check		OK	20:30
Air Blank		0.000	20:31
Control Test		0.078	20:31
Air Blank		0.000	20:32
Subject Sample	# 1	0.222	20:32
Breath Volume		1.898	20:32
Air Blank		0.000	20:33
Air Blank		0.000	20:35
Subject Sample	# 2	0.230	20:35
Breath Volume		1.753	20:35
Air Blank		0.000	20:36
Control Test		0.077	20:36
Air Blank		0.000	20:37
Diagnostics Check		OK	20:37

Data Download Date/Time: 05/16/2019 16:40

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINDER,BRONSON,M
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: PEREZ,NICHOLAS
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: AQUINO

DATE/TIME OF TEST: 04/28/2019 00:52:24

Diagnostics Check		OK	00:40	
Air Blank		0.000	00:41	
Control Test		0.077	00:41	
Air Blank		0.000	00:42	
Subject Sample	# 1	0.106	00:45	Volume Not Met
Breath Volume		0.074	00:45	
Air Blank		0.000	00:45	
Air Blank		0.000	00:47	
Subject Sample	# 2	0.108	00:50	Volume Not Met
Breath Volume		0.031	00:50	
Air Blank		0.000	00:51	
Control Test		0.078	00:51	
Air Blank		0.000	00:52	
Diagnostics Check		OK	00:52	

Data Download Date/Time: 05/16/2019 16:40

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINDER,BRONSON,M
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: PEREZ,NICHOLAS
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: AQUINO

DATE/TIME OF TEST: 04/28/2019 01:04:15

Diagnostics Check		OK	00:53	
Air Blank		0.000	00:54	
Control Test		0.076	00:54	
Air Blank		0.000	00:55	
Subject Sample	# 1	0.103	00:58	
Breath Volume		1.171	00:58	
Air Blank		0.000	00:58	
Air Blank		0.000	01:00	
Subject Sample	# 2	0.106	01:02	
Breath Volume		1.109	01:02	
Air Blank		0.000	01:03	
Control Test		0.074	01:03	Control Outside Tolerance
Air Blank		0.000	01:04	

Data Download Date/Time: 05/16/2019 16:40

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINDER,BRONSON,M
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: PEREZ,NICHOLAS
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: AQUINO

DATE/TIME OF TEST: 04/28/2019 01:13:47

Diagnostics Check		OK	01:06
Air Blank		0.000	01:06
Control Test		0.075	01:06
Air Blank		0.000	01:07
Subject Sample	# 1	0.110	01:09
Breath Volume		1.316	01:09
Air Blank		0.000	01:09
Air Blank		0.000	01:11
Subject Sample	# 2	0.112	01:12
Breath Volume		1.832	01:12
Air Blank		0.000	01:12
Control Test		0.076	01:13
Air Blank		0.000	01:13
Diagnostics Check		OK	01:13

Data Download Date/Time: 05/16/2019 16:40

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,ETHAN,C
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: BINDER,BRONSON
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: BURCH

DATE/TIME OF TEST: 04/28/2019 02:34:21

Diagnostics Check		OK	02:30	
Air Blank		0.000	02:30	
Control Test		0.076	02:31	
Air Blank		0.000	02:31	
Subject Sample	# 1		02:32	Subject Test Refused
Breath Volume		0.000	02:32	
Air Blank		0.000	02:33	
Control Test		0.076	02:33	
Air Blank		0.000	02:34	
Diagnostics Check		OK	02:34	

Data Download Date/Time: 05/16/2019 16:40

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINDER,BRONSON,M
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: BINDER,BRONSON
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: BURCH

DATE/TIME OF TEST: 04/28/2019 02:54:35

Diagnostics Check		OK	02:47
Air Blank		0.000	02:47
Control Test		0.078	02:48
Air Blank		0.000	02:48
Subject Sample	# 1	0.166	02:49
Breath Volume		1.597	02:49
Air Blank		0.000	02:50
Air Blank		0.000	02:52
Subject Sample	# 2	0.176	02:53
Breath Volume		1.714	02:53
Air Blank		0.000	02:53
Control Test		0.076	02:53
Air Blank		0.000	02:54
Diagnostics Check		OK	02:54

Data Download Date/Time: 05/16/2019 16:41

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALSH,SHANNON,L
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: WALSH,S
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 03/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A2
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: DISHNER

DATE/TIME OF TEST: 04/02/2019 22:11:47

Diagnostics Check		OK	22:01
Air Blank		0.000	22:02
Control Test		0.080	22:02
Air Blank		0.000	22:02
Subject Sample	# 1	0.137	22:04
Breath Volume		1.796	22:04
Air Blank		0.000	22:05
Air Blank		0.000	22:07
Subject Sample	# 2	0.126	22:10
Breath Volume		1.910	22:10
Air Blank		0.000	22:10
Control Test		0.080	22:11
Air Blank		0.000	22:11
Diagnostics Check		OK	22:11

Data Download Date/Time: 04/13/2019 10:46

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PINCUS,SETH,B
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: JOHNSON,STEVEN
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 03/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A2
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: VISCONTI

DATE/TIME OF TEST: 04/03/2019 18:34:25

Diagnostics Check		OK	18:30	
Air Blank		0.000	18:31	
Control Test		0.081	18:31	
Air Blank		0.000	18:32	
Subject Sample	# 1		18:33	Subject Test Refused
Breath Volume		0.000	18:33	
Air Blank		0.000	18:33	
Control Test		0.081	18:33	
Air Blank		0.000	18:34	
Diagnostics Check		OK	18:34	

Data Download Date/Time: 04/13/2019 10:46

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALSH,S,L
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: WALSH,S
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 04/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A2
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: LE

DATE/TIME OF TEST: 04/25/2019 03:24:10

Diagnostics Check		OK	03:17
Air Blank		0.000	03:18
Control Test		0.081	03:18
Air Blank		0.000	03:18
Subject Sample	# 1	0.159	03:19
Breath Volume		2.488	03:19
Air Blank		0.000	03:20
Air Blank		0.000	03:21
Subject Sample	# 2	0.155	03:22
Breath Volume		2.613	03:22
Air Blank		0.000	03:23
Control Test		0.080	03:23
Air Blank		0.000	03:24
Diagnostics Check		OK	03:24

Data Download Date/Time: 05/15/2019 15:17

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALSH,S,L
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: WALSH,S
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 04/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A2
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: LILES

DATE/TIME OF TEST: 04/26/2019 01:29:24

Diagnostics Check		OK	01:22
Air Blank		0.000	01:23
Control Test		0.080	01:23
Air Blank		0.000	01:23
Subject Sample	# 1	0.139	01:24
Breath Volume		2.121	01:24
Air Blank		0.000	01:25
Air Blank		0.000	01:27
Subject Sample	# 2	0.142	01:27
Breath Volume		2.464	01:27
Air Blank		0.000	01:28
Control Test		0.080	01:28
Air Blank		0.000	01:29
Diagnostics Check		OK	01:29

Data Download Date/Time: 05/15/2019 15:17

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,JONATHAN,B
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: RICHARD,LATASHA
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 04/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A2
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: KELLEY

DATE/TIME OF TEST: 04/27/2019 17:50:56

Diagnostics Check	OK	17:49	
Air Blank	0.000	17:50	
Control Test	0.080	17:50	RFI Detect
Air Blank	0.000	17:50	

Data Download Date/Time: 05/15/2019 15:17

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,JONATHAN,B
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: RICHARD,LATASHA
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 04/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A2
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: KELLEY

DATE/TIME OF TEST: 04/27/2019 17:59:09

Diagnostics Check		OK	17:52
Air Blank		0.000	17:53
Control Test		0.081	17:53
Air Blank		0.000	17:54
Subject Sample	# 1	0.166	17:54
Breath Volume		1.414	17:54
Air Blank		0.000	17:55
Air Blank		0.000	17:57
Subject Sample	# 2	0.165	17:57
Breath Volume		1.949	17:57
Air Blank		0.000	17:58
Control Test		0.080	17:58
Air Blank		0.000	17:59
Diagnostics Check		OK	17:59

Data Download Date/Time: 05/15/2019 15:17

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRINGTON,ALEX,M
BREATH TEST OPERATOR AGENCY: Belleview PD
ARRESTING OFFICER: HARRINGTON,ALEX
ARRESTING OFFICER AGENCY: Belleview PD
LAST AGENCY INSPECTION: 04/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 875088
DRY GAS EXPIRY DATE: 12/12/2020
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: GRAZIANI

DATE/TIME OF TEST: 04/21/2019 02:21:07

Diagnostics Check		OK	02:14
Air Blank		0.000	02:14
Control Test		0.079	02:15
Air Blank		0.000	02:15
Subject Sample	# 1	0.153	02:16
Breath Volume		1.214	02:16
Air Blank		0.000	02:17
Air Blank		0.000	02:19
Subject Sample	# 2	0.158	02:19
Breath Volume		1.652	02:19
Air Blank		0.000	02:20
Control Test		0.078	02:20
Air Blank		0.000	02:21
Diagnostics Check		OK	02:21

Data Download Date/Time: 05/22/2019 16:25

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STILLION,JUSTIN,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: STILLION,JUSTIN
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 04/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 875088
DRY GAS EXPIRY DATE: 12/12/2020
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: QUINN

DATE/TIME OF TEST: 04/25/2019 00:38:10

Diagnostics Check		OK	00:35	
Air Blank		0.000	00:35	
Control Test		0.079	00:36	
Air Blank		0.000	00:36	
Subject Sample	# 1		00:36	Subject Test Refused
Breath Volume		0.000	00:36	
Air Blank		0.000	00:37	
Control Test		0.078	00:37	
Air Blank		0.000	00:38	
Diagnostics Check		OK	00:38	

Data Download Date/Time: 05/22/2019 16:25

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAINES,ELDON,L
BREATH TEST OPERATOR AGENCY: Dunnellon PD
ARRESTING OFFICER: RAINES,ELDON
ARRESTING OFFICER AGENCY: Dunnellon PD
LAST AGENCY INSPECTION: 04/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 875088
DRY GAS EXPIRY DATE: 12/12/2020
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: VARTY

DATE/TIME OF TEST: 04/26/2019 02:28:34

Diagnostics Check		OK	02:22
Air Blank		0.000	02:22
Control Test		0.078	02:23
Air Blank		0.000	02:23
Subject Sample	# 1	0.209	02:24
Breath Volume		2.664	02:24
Air Blank		0.000	02:24
Air Blank		0.000	02:26
Subject Sample	# 2	0.198	02:27
Breath Volume		2.542	02:27
Air Blank		0.000	02:27
Control Test		0.077	02:27
Air Blank		0.000	02:28
Diagnostics Check		OK	02:28

Data Download Date/Time: 05/22/2019 16:25

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: PAYNTER,H
ARRESTING OFFICER AGENCY: FHP TROOP C
LAST AGENCY INSPECTION: 03/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: ARROYO JR

DATE/TIME OF TEST: 04/02/2019 03:45:32

Diagnostics Check		OK	03:39
Air Blank		0.000	03:39
Control Test		0.080	03:40
Air Blank		0.000	03:40
Subject Sample	# 1	0.128	03:41
Breath Volume		3.746	03:41
Air Blank		0.000	03:41
Air Blank		0.000	03:43
Subject Sample	# 2	0.128	03:44
Breath Volume		3.765	03:44
Air Blank		0.000	03:44
Control Test		0.079	03:44
Air Blank		0.000	03:45
Diagnostics Check		OK	03:45

Data Download Date/Time: 04/04/2019 17:10

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: VAN TREESE,S
ARRESTING OFFICER AGENCY: TAMPA PD
LAST AGENCY INSPECTION: 03/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:03

SUBJECT'S LAST NAME: GLEASON

DATE/TIME OF TEST: 04/03/2019 22:30:05

Diagnostics Check		OK	22:23
Air Blank		0.000	22:23
Control Test		0.079	22:24
Air Blank		0.000	22:24
Subject Sample	# 1	0.372	22:25
Breath Volume		2.710	22:25
Air Blank		0.000	22:26
Air Blank		0.000	22:27
Subject Sample	# 2	0.370	22:28
Breath Volume		2.292	22:28
Air Blank		0.000	22:29
Control Test		0.078	22:29
Air Blank		0.000	22:30
Diagnostics Check		OK	22:30

Data Download Date/Time: 04/04/2019 17:10

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: LYON,M
ARRESTING OFFICER AGENCY: TAMPA PD
LAST AGENCY INSPECTION: 03/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 04/04/2019 00:10:54

Diagnostics Check		OK	00:03
Air Blank		0.000	00:04
Control Test		0.079	00:04
Air Blank		0.000	00:05
Subject Sample	# 1	0.182	00:06
Breath Volume		3.378	00:06
Air Blank		0.000	00:06
Air Blank		0.000	00:08
Subject Sample	# 2	0.185	00:09
Breath Volume		2.441	00:09
Air Blank		0.000	00:09
Control Test		0.079	00:10
Air Blank		0.000	00:10
Diagnostics Check		OK	00:10

Data Download Date/Time: 04/04/2019 17:10

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES AVERY,D,C
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: GONZALEZ,D
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 03/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: KINSEY

DATE/TIME OF TEST: 04/04/2019 01:33:40

Diagnostics Check		OK	01:27
Air Blank		0.000	01:27
Control Test		0.080	01:28
Air Blank		0.000	01:28
Subject Sample	# 1	0.020	01:29
Breath Volume		2.187	01:29
Air Blank		0.000	01:29
Air Blank		0.000	01:31
Subject Sample	# 2	0.020	01:32
Breath Volume		1.769	01:32
Air Blank		0.000	01:32
Control Test		0.079	01:33
Air Blank		0.000	01:33
Diagnostics Check		OK	01:33

Data Download Date/Time: 04/04/2019 17:10

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: CAMPBELL,J
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 03/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:12

SUBJECT'S LAST NAME: WASIM

DATE/TIME OF TEST: 04/04/2019 05:41:42

Diagnostics Check		OK	05:32
Air Blank		0.000	05:32
Control Test		0.079	05:33
Air Blank		0.000	05:33
Subject Sample	# 1	0.137	05:34
Breath Volume		1.503	05:34
Air Blank		0.000	05:34
Air Blank		0.000	05:36
Subject Sample	# 2	0.163	05:37
Breath Volume		3.304	05:37
Air Blank		0.000	05:37
Air Blank		0.000	05:39
Subject Sample	# 3	0.155	05:40
Breath Volume		3.425	05:40
Air Blank		0.000	05:40
Control Test		0.080	05:41
Air Blank		0.000	05:41
Diagnostics Check		OK	05:41

Data Download Date/Time: 04/04/2019 17:10

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: TAMPA PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: TATE

DATE/TIME OF TEST: 04/06/2019 04:10:16

Diagnostics Check		OK	04:03
Air Blank		0.000	04:04
Control Test		0.080	04:04
Air Blank		0.000	04:05
Subject Sample	# 1	0.102	04:05
Breath Volume		3.117	04:05
Air Blank		0.000	04:06
Air Blank		0.000	04:08
Subject Sample	# 2	0.098	04:08
Breath Volume		2.769	04:08
Air Blank		0.000	04:09
Control Test		0.079	04:09
Air Blank		0.000	04:10
Diagnostics Check		OK	04:10

Data Download Date/Time: 05/06/2019 13:42

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES AVERY,D,C
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: SIMPKINS,D
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: VELEZ

DATE/TIME OF TEST: 04/06/2019 22:21:11

Diagnostics Check		OK	22:15
Air Blank		0.000	22:15
Control Test		0.080	22:15
Air Blank		0.000	22:16
Subject Sample	# 1	0.217	22:16
Breath Volume		2.277	22:16
Air Blank		0.000	22:17
Air Blank		0.000	22:19
Subject Sample	# 2	0.217	22:19
Breath Volume		1.609	22:19
Air Blank		0.000	22:20
Control Test		0.079	22:20
Air Blank		0.000	22:21
Diagnostics Check		OK	22:21

Data Download Date/Time: 05/06/2019 13:42

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: THORNE,S
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: WAITE

DATE/TIME OF TEST: 04/07/2019 03:14:14

Diagnostics Check		OK	03:07
Air Blank		0.000	03:08
Control Test		0.079	03:08
Air Blank		0.000	03:09
Subject Sample	# 1	0.090	03:09
Breath Volume		1.816	03:09
Air Blank		0.000	03:10
Air Blank		0.000	03:12
Subject Sample	# 2	0.096	03:12
Breath Volume		2.195	03:12
Air Blank		0.000	03:13
Control Test		0.079	03:13
Air Blank		0.000	03:14
Diagnostics Check		OK	03:14

Data Download Date/Time: 05/06/2019 13:42

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES AVERY,D,C
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: TAMPA PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: WALIGA

DATE/TIME OF TEST: 04/07/2019 22:01:46

Diagnostics Check		OK	21:55
Air Blank		0.000	21:56
Control Test		0.080	21:56
Air Blank		0.000	21:57
Subject Sample	# 1	0.000	21:57
Breath Volume		4.027	21:57
Air Blank		0.000	21:58
Air Blank		0.000	22:00
Subject Sample	# 2	0.000	22:00
Breath Volume		3.093	22:00
Air Blank		0.000	22:00
Control Test		0.080	22:01
Air Blank		0.000	22:01
Diagnostics Check		OK	22:01

Data Download Date/Time: 05/06/2019 13:42

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: DRISCOLL,B
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: FELIX

DATE/TIME OF TEST: 04/08/2019 03:27:35

Diagnostics Check		OK	03:13	
Air Blank		0.000	03:13	
Control Test		0.081	03:13	
Air Blank		0.000	03:14	
Subject Sample	# 1	0.112	03:17	Volume Not Met
Breath Volume		0.417	03:17	
Air Blank		0.000	03:18	
Air Blank		0.000	03:20	
Subject Sample	# 2	0.112	03:22	
Breath Volume		1.238	03:22	
Air Blank		0.000	03:22	
Air Blank		0.000	03:24	
Subject Sample	# 3	0.086	03:26	
Breath Volume		1.800	03:26	
Air Blank		0.000	03:26	
Control Test		0.050	03:27	Control Outside Tolerance
Air Blank		0.000	03:27	

Data Download Date/Time: 05/06/2019 13:42

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: DRISCOLL,B
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: FELIX

DATE/TIME OF TEST: 04/08/2019 03:45:46

Diagnostics Check		OK	03:30	
Air Blank		0.000	03:30	
Control Test		0.079	03:30	
Air Blank		0.000	03:31	
Subject Sample	# 1	0.103	03:34	Volume Not Met
Breath Volume		0.300	03:34	
Air Blank		0.000	03:35	
Air Blank		0.000	03:36	
Subject Sample	# 2	0.120	03:38	
Breath Volume		1.152	03:38	
Air Blank		0.000	03:39	
Air Blank		0.000	03:41	
Subject Sample	# 3	0.091	03:44	Volume Not Met
Breath Volume		0.066	03:44	
Air Blank		0.000	03:44	
Control Test		0.079	03:45	
Air Blank		0.000	03:45	
Diagnostics Check		OK	03:45	

Data Download Date/Time: 05/06/2019 13:42

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: TAMPA PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: LUNA VICTORIA

DATE/TIME OF TEST: 04/11/2019 02:27:44

Diagnostics Check		OK	02:20
Air Blank		0.000	02:21
Control Test		0.080	02:21
Air Blank		0.000	02:21
Subject Sample	# 1	0.128	02:23
Breath Volume		1.218	02:23
Air Blank		0.000	02:24
Air Blank		0.000	02:25
Subject Sample	# 2	0.123	02:26
Breath Volume		1.363	02:26
Air Blank		0.000	02:26
Control Test		0.078	02:27
Air Blank		0.000	02:27
Diagnostics Check		OK	02:27

Data Download Date/Time: 05/06/2019 13:42

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: ANDREWS,S
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 04/12/2019 02:47:59

Diagnostics Check		OK	02:41
Air Blank		0.000	02:42
Control Test		0.079	02:42
Air Blank		0.000	02:42
Subject Sample	# 1	0.148	02:43
Breath Volume		1.769	02:43
Air Blank		0.000	02:44
Air Blank		0.000	02:45
Subject Sample	# 2	0.149	02:46
Breath Volume		1.707	02:46
Air Blank		0.000	02:47
Control Test		0.078	02:47
Air Blank		0.000	02:47
Diagnostics Check		OK	02:47

Data Download Date/Time: 05/06/2019 13:42

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: THORNE,S
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: SEXTON

DATE/TIME OF TEST: 04/12/2019 23:57:04

Diagnostics Check		OK	23:50
Air Blank		0.000	23:51
Control Test		0.079	23:51
Air Blank		0.000	23:51
Subject Sample	# 1	0.206	23:52
Breath Volume		1.839	23:52
Air Blank		0.000	23:53
Air Blank		0.000	23:55
Subject Sample	# 2	0.213	23:55
Breath Volume		2.105	23:55
Air Blank		0.000	23:56
Control Test		0.079	23:56
Air Blank		0.000	23:56
Diagnostics Check		OK	23:57

Data Download Date/Time: 05/06/2019 13:42

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: GONZALEZ,D
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: PAREE

DATE/TIME OF TEST: 04/13/2019 03:27:11

Diagnostics Check		OK	03:20
Air Blank		0.000	03:21
Control Test		0.080	03:21
Air Blank		0.000	03:22
Subject Sample	# 1	0.148	03:22
Breath Volume		2.054	03:22
Air Blank		0.000	03:23
Air Blank		0.000	03:25
Subject Sample	# 2	0.149	03:25
Breath Volume		2.777	03:25
Air Blank		0.000	03:26
Control Test		0.078	03:26
Air Blank		0.000	03:27
Diagnostics Check		OK	03:27

Data Download Date/Time: 05/06/2019 13:42

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: MCDONALD

DATE/TIME OF TEST: 04/13/2019 04:51:15

Diagnostics Check		OK	04:39	
Air Blank		0.000	04:40	
Control Test		0.080	04:40	
Air Blank		0.000	04:40	
Subject Sample	# 1	0.118	04:44	Volume Not Met
Breath Volume		0.468	04:44	
Air Blank		0.000	04:44	
Air Blank		0.000	04:46	
Subject Sample	# 2	0.106	04:49	Volume Not Met
Breath Volume		0.265	04:49	
Air Blank		0.000	04:50	
Control Test		0.078	04:50	
Air Blank		0.000	04:51	
Diagnostics Check		OK	04:51	

Data Download Date/Time: 05/06/2019 13:42

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: TAMPA PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: PIRTLE

DATE/TIME OF TEST: 04/16/2019 02:13:34

Diagnostics Check		OK	02:06
Air Blank		0.000	02:06
Control Test		0.080	02:07
Air Blank		0.000	02:07
Subject Sample	# 1	0.175	02:08
Breath Volume		1.562	02:08
Air Blank		0.000	02:09
Air Blank		0.000	02:10
Subject Sample	# 2	0.170	02:12
Breath Volume		1.617	02:12
Air Blank		0.000	02:12
Control Test		0.079	02:12
Air Blank		0.000	02:13
Diagnostics Check		OK	02:13

Data Download Date/Time: 05/06/2019 13:42

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES AVERY,D,C
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: SULAIMAN,R
ARRESTING OFFICER AGENCY: FHP TROOP C
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:17

SUBJECT'S LAST NAME: DEMERASKI

DATE/TIME OF TEST: 04/16/2019 19:43:41

Diagnostics Check		OK	19:37
Air Blank		0.000	19:38
Control Test		0.080	19:38
Air Blank		0.000	19:39
Subject Sample	# 1	0.127	19:39
Breath Volume		3.898	19:39
Air Blank		0.000	19:39
Air Blank		0.000	19:41
Subject Sample	# 2	0.119	19:42
Breath Volume		4.246	19:42
Air Blank		0.000	19:42
Control Test		0.079	19:43
Air Blank		0.000	19:43
Diagnostics Check		OK	19:43

Data Download Date/Time: 05/06/2019 13:42

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: DRISCOLL,B
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: BISHOP-JONES

DATE/TIME OF TEST: 04/20/2019 00:43:55

Diagnostics Check		OK	00:37
Air Blank		0.000	00:38
Control Test		0.080	00:38
Air Blank		0.000	00:38
Subject Sample	# 1	0.169	00:39
Breath Volume		2.507	00:39
Air Blank		0.000	00:39
Air Blank		0.000	00:41
Subject Sample	# 2	0.165	00:42
Breath Volume		1.933	00:42
Air Blank		0.000	00:43
Control Test		0.079	00:43
Air Blank		0.000	00:43
Diagnostics Check		OK	00:43

Data Download Date/Time: 05/06/2019 13:42

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: DOBSON,B
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: HUGHES

DATE/TIME OF TEST: 04/20/2019 03:57:22

Diagnostics Check		OK	03:50
Air Blank		0.000	03:50
Control Test		0.080	03:51
Air Blank		0.000	03:51
Subject Sample	# 1	0.168	03:52
Breath Volume		1.285	03:52
Air Blank		0.000	03:53
Air Blank		0.000	03:55
Subject Sample	# 2	0.183	03:55
Breath Volume		2.136	03:55
Air Blank		0.000	03:56
Control Test		0.079	03:56
Air Blank		0.000	03:57
Diagnostics Check		OK	03:57

Data Download Date/Time: 05/06/2019 13:42

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES AVERY,D,C
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: RUIZ

DATE/TIME OF TEST: 04/21/2019 00:37:00

Diagnostics Check		OK	00:31
Air Blank		0.000	00:31
Control Test		0.080	00:32
Air Blank		0.000	00:32
Subject Sample	# 1	0.000	00:32
Breath Volume		1.820	00:32
Air Blank		0.000	00:33
Air Blank		0.000	00:35
Subject Sample	# 2	0.000	00:35
Breath Volume		1.933	00:35
Air Blank		0.000	00:36
Control Test		0.079	00:36
Air Blank		0.000	00:36
Diagnostics Check		OK	00:37

Data Download Date/Time: 05/06/2019 13:42

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: DRISCOLL,B
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: PERSAUD

DATE/TIME OF TEST: 04/21/2019 02:11:44

Diagnostics Check		OK	01:56	
Air Blank		0.000	01:57	
Control Test		0.080	01:57	
Air Blank		0.000	01:58	
Subject Sample	# 1	0.265	02:00	
Breath Volume		1.281	02:00	
Air Blank		0.000	02:00	
Air Blank		0.000	02:02	
Subject Sample	# 2	0.257	02:05	Volume Not Met
Breath Volume		0.167	02:05	
Air Blank		0.000	02:06	
Air Blank		0.000	02:08	
Subject Sample	# 3	0.270	02:10	
Breath Volume		2.070	02:10	
Air Blank		0.000	02:10	
Control Test		0.078	02:11	
Air Blank		0.000	02:11	
Diagnostics Check		OK	02:11	

Data Download Date/Time: 05/06/2019 13:42

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: PAYNTER,H
ARRESTING OFFICER AGENCY: FHP TROOP C
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: RODRIQUEZ JR

DATE/TIME OF TEST: 04/21/2019 04:56:45

Diagnostics Check		OK	04:49
Air Blank		0.000	04:50
Control Test		0.079	04:50
Air Blank		0.000	04:51
Subject Sample	# 1	0.116	04:52
Breath Volume		1.593	04:52
Air Blank		0.000	04:52
Air Blank		0.000	04:54
Subject Sample	# 2	0.114	04:55
Breath Volume		1.730	04:55
Air Blank		0.000	04:55
Control Test		0.080	04:56
Air Blank		0.000	04:56
Diagnostics Check		OK	04:56

Data Download Date/Time: 05/06/2019 13:42

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: ARCHAMBEAU,M
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:43

SUBJECT'S LAST NAME: VLAZNY

DATE/TIME OF TEST: 04/21/2019 22:10:22

Diagnostics Check		OK	22:03
Air Blank		0.000	22:04
Control Test		0.080	22:04
Air Blank		0.000	22:05
Subject Sample	# 1	0.169	22:05
Breath Volume		3.285	22:05
Air Blank		0.000	22:06
Air Blank		0.000	22:08
Subject Sample	# 2	0.172	22:08
Breath Volume		2.507	22:08
Air Blank		0.000	22:09
Control Test		0.080	22:09
Air Blank		0.000	22:10
Diagnostics Check		OK	22:10

Data Download Date/Time: 05/06/2019 13:42

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: ORNDORFF,W
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: VELEZ

DATE/TIME OF TEST: 04/27/2019 21:06:45

Diagnostics Check		OK	21:00
Air Blank		0.000	21:00
Control Test		0.081	21:01
Air Blank		0.000	21:01
Subject Sample	# 1	0.288	21:02
Breath Volume		1.621	21:02
Air Blank		0.000	21:02
Air Blank		0.000	21:04
Subject Sample	# 2	0.274	21:05
Breath Volume		2.046	21:05
Air Blank		0.000	21:05
Control Test		0.079	21:06
Air Blank		0.000	21:06
Diagnostics Check		OK	21:06

Data Download Date/Time: 05/06/2019 13:42

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: DE JESUS,V
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: VAZQUEZ HERNANDEZ

DATE/TIME OF TEST: 04/28/2019 01:07:53

Diagnostics Check		OK	01:01
Air Blank		0.000	01:01
Control Test		0.079	01:01
Air Blank		0.000	01:02
Subject Sample	# 1	0.129	01:03
Breath Volume		2.398	01:03
Air Blank		0.000	01:04
Air Blank		0.000	01:05
Subject Sample	# 2	0.129	01:06
Breath Volume		2.312	01:06
Air Blank		0.000	01:06
Control Test		0.080	01:07
Air Blank		0.000	01:07
Diagnostics Check		OK	01:07

Data Download Date/Time: 05/06/2019 13:42

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: TAMPA PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: AMBROSIO

DATE/TIME OF TEST: 04/28/2019 03:58:32

Diagnostics Check		OK	03:52
Air Blank		0.000	03:52
Control Test		0.079	03:53
Air Blank		0.000	03:53
Subject Sample	# 1	0.104	03:54
Breath Volume		1.781	03:54
Air Blank		0.000	03:54
Air Blank		0.000	03:56
Subject Sample	# 2	0.108	03:57
Breath Volume		1.539	03:57
Air Blank		0.000	03:57
Control Test		0.079	03:57
Air Blank		0.000	03:58
Diagnostics Check		OK	03:58

Data Download Date/Time: 05/06/2019 13:42

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 04/30/2019 03:38:26

Diagnostics Check		OK	03:33	
Air Blank		0.000	03:34	
Control Test		0.081	03:34	
Air Blank		0.000	03:35	
Subject Sample	# 1		03:37	Subject Test Refused
Breath Volume		0.000	03:37	
Air Blank		0.000	03:37	
Control Test		0.080	03:37	
Air Blank		0.000	03:38	
Diagnostics Check		OK	03:38	

Data Download Date/Time: 05/06/2019 13:42

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES AVERY,D,C
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: HALL,L
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: BIRGE

DATE/TIME OF TEST: 04/30/2019 21:34:18

Diagnostics Check		OK	21:26
Air Blank		0.000	21:26
Control Test		0.080	21:26
Air Blank		0.000	21:27
Subject Sample	# 1	0.151	21:29
Breath Volume		1.531	21:29
Air Blank		0.000	21:29
Air Blank		0.000	21:31
Subject Sample	# 2	0.151	21:32
Breath Volume		2.140	21:32
Air Blank		0.000	21:33
Control Test		0.079	21:33
Air Blank		0.000	21:34
Diagnostics Check		OK	21:34

Data Download Date/Time: 05/06/2019 13:42

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERRA,CORP,
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: SERRA,CORP
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:26

SUBJECT'S LAST NAME: HESSEE

DATE/TIME OF TEST: 04/04/2019 21:09:23

Diagnostics Check		OK	21:02
Air Blank		0.000	21:03
Control Test		0.077	21:03
Air Blank		0.000	21:04
Subject Sample	# 1	0.152	21:04
Breath Volume		2.222	21:04
Air Blank		0.000	21:05
Air Blank		0.000	21:07
Subject Sample	# 2	0.154	21:07
Breath Volume		2.109	21:07
Air Blank		0.000	21:08
Control Test		0.076	21:08
Air Blank		0.000	21:09
Diagnostics Check		OK	21:09

Data Download Date/Time: 04/26/2019 01:18

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREGORY,AMANDA,M
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: GREGORY,AMANDA
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 04/05/2019 01:20:37

Diagnostics Check		OK	01:13
Air Blank		0.000	01:13
Control Test		0.077	01:13
Air Blank		0.000	01:14
Subject Sample	# 1	0.158	01:14
Breath Volume		1.386	01:14
Air Blank		0.000	01:15
Air Blank		0.000	01:17
Subject Sample	# 2	0.156	01:18
Breath Volume		1.664	01:18
Air Blank		0.000	01:19
Control Test		0.077	01:19
Air Blank		0.000	01:20
Diagnostics Check		OK	01:20

Data Download Date/Time: 04/26/2019 01:18

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLASZYK,JACOB,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BLASZYK,JACOB
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 04/07/2019 02:36:25

Diagnostics Check		OK	02:30
Air Blank		0.000	02:30
Control Test		0.078	02:30
Air Blank		0.000	02:31
Subject Sample	# 1	0.054	02:31
Breath Volume		1.117	02:31
Air Blank		0.000	02:32
Air Blank		0.000	02:34
Subject Sample	# 2	0.055	02:34
Breath Volume		1.976	02:34
Air Blank		0.000	02:35
Control Test		0.077	02:35
Air Blank		0.000	02:36
Diagnostics Check		OK	02:36

Data Download Date/Time: 04/26/2019 01:18

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:28

SUBJECT'S LAST NAME: MARTIN-CHILEL

DATE/TIME OF TEST: 04/07/2019 20:17:49

Diagnostics Check		OK	20:10
Air Blank		0.000	20:11
Control Test		0.076	20:11
Air Blank		0.000	20:11
Subject Sample	# 1	0.247	20:12
Breath Volume		1.457	20:12
Air Blank		0.000	20:13
Air Blank		0.000	20:15
Subject Sample	# 2	0.255	20:16
Breath Volume		1.679	20:16
Air Blank		0.000	20:16
Control Test		0.076	20:17
Air Blank		0.000	20:17
Diagnostics Check		OK	20:17

Data Download Date/Time: 04/26/2019 01:18

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BREVIK,LONNIE,J
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: BREVIK,LONNIE
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: SYDNAM

DATE/TIME OF TEST: 04/07/2019 23:50:14

Diagnostics Check		OK	23:43
Air Blank		0.000	23:44
Control Test		0.077	23:44
Air Blank		0.000	23:45
Subject Sample	# 1	0.098	23:45
Breath Volume		1.460	23:45
Air Blank		0.000	23:46
Air Blank		0.000	23:48
Subject Sample	# 2	0.096	23:48
Breath Volume		2.003	23:48
Air Blank		0.000	23:49
Control Test		0.077	23:49
Air Blank		0.000	23:50
Diagnostics Check		OK	23:50

Data Download Date/Time: 04/26/2019 01:18

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLASZYK,JACOB,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BLASZYK,JACOB
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: HENRICHSEN

DATE/TIME OF TEST: 04/11/2019 01:53:04

Diagnostics Check		OK	01:46
Air Blank		0.000	01:46
Control Test		0.076	01:46
Air Blank		0.000	01:47
Subject Sample	# 1	0.169	01:48
Breath Volume		1.136	01:48
Air Blank		0.000	01:49
Air Blank		0.000	01:50
Subject Sample	# 2	0.176	01:51
Breath Volume		1.933	01:51
Air Blank		0.000	01:52
Control Test		0.076	01:52
Air Blank		0.000	01:52
Diagnostics Check		OK	01:53

Data Download Date/Time: 04/26/2019 01:18

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CERNUTO,MATTHEW,D
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: CERNUTO,MATTHEW
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 12:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:26

SUBJECT'S LAST NAME: STIGLICH

DATE/TIME OF TEST: 04/11/2019 13:23:28

Diagnostics Check		OK	13:17
Air Blank		0.000	13:18
Control Test		0.077	13:18
Air Blank		0.000	13:18
Subject Sample	# 1	0.000	13:19
Breath Volume		1.667	13:19
Air Blank		0.000	13:19
Air Blank		0.000	13:21
Subject Sample	# 2	0.000	13:22
Breath Volume		1.722	13:22
Air Blank		0.000	13:22
Control Test		0.077	13:22
Air Blank		0.000	13:23
Diagnostics Check		OK	13:23

Data Download Date/Time: 04/26/2019 01:18

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLASZYK,JACOB,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BLASZYK,JACOB
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:02

SUBJECT'S LAST NAME: PARLIER

DATE/TIME OF TEST: 04/11/2019 21:29:15

Diagnostics Check		OK	21:22
Air Blank		0.000	21:22
Control Test		0.077	21:23
Air Blank		0.000	21:23
Subject Sample	# 1	0.236	21:24
Breath Volume		1.863	21:24
Air Blank		0.000	21:25
Air Blank		0.000	21:26
Subject Sample	# 2	0.227	21:27
Breath Volume		2.085	21:27
Air Blank		0.000	21:28
Control Test		0.077	21:28
Air Blank		0.000	21:29
Diagnostics Check		OK	21:29

Data Download Date/Time: 04/26/2019 01:18

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEEDHAM,K,
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: PALMIERI,A
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:27

SUBJECT'S LAST NAME: VEGA

DATE/TIME OF TEST: 04/12/2019 20:54:26

Diagnostics Check		OK	20:48
Air Blank		0.000	20:48
Control Test		0.078	20:48
Air Blank		0.000	20:49
Subject Sample	# 1	0.198	20:49
Breath Volume		2.269	20:49
Air Blank		0.000	20:50
Air Blank		0.000	20:52
Subject Sample	# 2	0.207	20:52
Breath Volume		2.187	20:52
Air Blank		0.000	20:53
Control Test		0.077	20:53
Air Blank		0.000	20:54
Diagnostics Check		OK	20:54

Data Download Date/Time: 04/26/2019 01:18

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: DEAN

DATE/TIME OF TEST: 04/14/2019 03:07:52

Diagnostics Check		OK	02:59
Air Blank		0.000	03:00
Control Test		0.078	03:00
Air Blank		0.000	03:01
Subject Sample	# 1	0.181	03:02
Breath Volume		1.160	03:02
Air Blank		0.000	03:03
Air Blank		0.000	03:05
Subject Sample	# 2	0.176	03:06
Breath Volume		1.894	03:06
Air Blank		0.000	03:06
Control Test		0.076	03:07
Air Blank		0.000	03:07
Diagnostics Check		OK	03:07

Data Download Date/Time: 04/26/2019 01:18

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: GARCIA,NELSON
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: ZAPETA TZUL

DATE/TIME OF TEST: 04/14/2019 04:02:55

Diagnostics Check		OK	03:55
Air Blank		0.000	03:56
Control Test		0.078	03:56
Air Blank		0.000	03:57
Subject Sample	# 1	0.113	03:58
Breath Volume		2.128	03:58
Air Blank		0.000	03:59
Air Blank		0.000	04:00
Subject Sample	# 2	0.105	04:01
Breath Volume		1.292	04:01
Air Blank		0.000	04:01
Control Test		0.077	04:02
Air Blank		0.000	04:02
Diagnostics Check		OK	04:02

Data Download Date/Time: 04/26/2019 01:18

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ISHAM,MATTHEW,H
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: ISHAM,MATTHEW
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:17

SUBJECT'S LAST NAME: LEONARD

DATE/TIME OF TEST: 04/14/2019 20:45:33

Diagnostics Check		OK	20:39
Air Blank		0.000	20:39
Control Test		0.077	20:40
Air Blank		0.000	20:40
Subject Sample	# 1	0.220	20:40
Breath Volume		2.890	20:40
Air Blank		0.000	20:41
Air Blank		0.000	20:43
Subject Sample	# 2	0.217	20:43
Breath Volume		3.082	20:43
Air Blank		0.000	20:44
Control Test		0.076	20:44
Air Blank		0.000	20:45
Diagnostics Check		OK	20:45

Data Download Date/Time: 04/26/2019 01:18

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RICKY,
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: PEREZ,RICKY
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: POULIN

DATE/TIME OF TEST: 04/15/2019 23:47:21

Diagnostics Check		OK	23:41
Air Blank		0.000	23:41
Control Test		0.077	23:42
Air Blank		0.000	23:42
Subject Sample	# 1	0.000	23:43
Breath Volume		1.351	23:43
Air Blank		0.000	23:43
Air Blank		0.000	23:45
Subject Sample	# 2	0.000	23:45
Breath Volume		1.156	23:45
Air Blank		0.000	23:46
Control Test		0.077	23:46
Air Blank		0.000	23:47
Diagnostics Check		OK	23:47

Data Download Date/Time: 04/26/2019 01:18

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPLAN,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: CAPLAN,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: JOHANSEN

DATE/TIME OF TEST: 04/17/2019 01:55:33

Diagnostics Check		OK	01:48
Air Blank		0.000	01:49
Control Test		0.077	01:49
Air Blank		0.000	01:50
Subject Sample	# 1	0.273	01:50
Breath Volume		2.675	01:50
Air Blank		0.000	01:51
Air Blank		0.000	01:53
Subject Sample	# 2	0.291	01:53
Breath Volume		2.257	01:53
Air Blank		0.000	01:54
Control Test		0.076	01:54
Air Blank		0.000	01:55
Diagnostics Check		OK	01:55

Data Download Date/Time: 04/26/2019 01:18

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: MONROIG,ANIBAL
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: MCKEE

DATE/TIME OF TEST: 04/19/2019 01:37:20

Diagnostics Check		OK	01:31
Air Blank		0.000	01:31
Control Test		0.077	01:31
Air Blank		0.000	01:32
Subject Sample	# 1	0.209	01:32
Breath Volume		2.601	01:32
Air Blank		0.000	01:33
Air Blank		0.000	01:35
Subject Sample	# 2	0.203	01:35
Breath Volume		2.433	01:35
Air Blank		0.000	01:36
Control Test		0.078	01:36
Air Blank		0.000	01:37
Diagnostics Check		OK	01:37

Data Download Date/Time: 04/26/2019 01:18

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: RITTER

DATE/TIME OF TEST: 04/20/2019 00:41:12

Diagnostics Check		OK	00:34
Air Blank		0.000	00:35
Control Test		0.077	00:35
Air Blank		0.000	00:36
Subject Sample	# 1	0.110	00:36
Breath Volume		3.957	00:36
Air Blank		0.000	00:37
Air Blank		0.000	00:39
Subject Sample	# 2	0.112	00:39
Breath Volume		3.964	00:39
Air Blank		0.000	00:40
Control Test		0.078	00:40
Air Blank		0.000	00:41
Diagnostics Check		OK	00:41

Data Download Date/Time: 04/26/2019 01:18

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: MEJIA-SOLIS

DATE/TIME OF TEST: 04/20/2019 04:03:15

Diagnostics Check		OK	03:54
Air Blank		0.000	03:55
Control Test		0.078	03:55
Air Blank		0.000	03:56
Subject Sample	# 1	0.279	03:57
Breath Volume		1.496	03:57
Air Blank		0.000	03:57
Air Blank		0.000	03:59
Subject Sample	# 2	0.272	04:01
Breath Volume		1.539	04:01
Air Blank		0.000	04:02
Control Test		0.076	04:02
Air Blank		0.000	04:03
Diagnostics Check		OK	04:03

Data Download Date/Time: 04/26/2019 01:18

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLIZZI,CHRISTINE,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: POLIZZI,CHRISTINE
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: BRENNAN

DATE/TIME OF TEST: 04/21/2019 00:33:02

Diagnostics Check		OK	00:26
Air Blank		0.000	00:27
Control Test		0.077	00:27
Air Blank		0.000	00:28
Subject Sample	# 1	0.203	00:28
Breath Volume		2.156	00:28
Air Blank		0.000	00:29
Air Blank		0.000	00:31
Subject Sample	# 2	0.201	00:31
Breath Volume		2.167	00:31
Air Blank		0.000	00:32
Control Test		0.077	00:32
Air Blank		0.000	00:32
Diagnostics Check		OK	00:33

Data Download Date/Time: 04/26/2019 01:18

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEEDHAM,K,
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: PALMIERI,A
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:03

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 04/21/2019 22:31:12

Diagnostics Check		OK	22:24
Air Blank		0.000	22:25
Control Test		0.076	22:25
Air Blank		0.000	22:26
Subject Sample	# 1	0.139	22:26
Breath Volume		2.308	22:26
Air Blank		0.000	22:27
Air Blank		0.000	22:28
Subject Sample	# 2	0.135	22:29
Breath Volume		2.226	22:29
Air Blank		0.000	22:30
Control Test		0.077	22:30
Air Blank		0.000	22:31
Diagnostics Check		OK	22:31

Data Download Date/Time: 04/26/2019 01:18

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPLAN,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: CAPLAN,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: FOCHT

DATE/TIME OF TEST: 04/22/2019 00:07:03

Diagnostics Check		OK	00:00
Air Blank		0.000	00:00
Control Test		0.078	00:01
Air Blank		0.000	00:01
Subject Sample	# 1	0.233	00:02
Breath Volume		1.949	00:02
Air Blank		0.000	00:02
Air Blank		0.000	00:04
Subject Sample	# 2	0.230	00:05
Breath Volume		1.894	00:05
Air Blank		0.000	00:06
Control Test		0.076	00:06
Air Blank		0.000	00:06
Diagnostics Check		OK	00:07

Data Download Date/Time: 04/26/2019 01:18

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENOVIL,GEDEON,
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BRENOVIL,GEDEON
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 14:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:19

SUBJECT'S LAST NAME: MOSHER

DATE/TIME OF TEST: 04/22/2019 15:47:59

Diagnostics Check		OK	15:41	
Air Blank		0.000	15:41	
Control Test		0.077	15:42	
Air Blank		0.000	15:42	
Subject Sample	# 1	0.287	15:43	
Breath Volume		2.042	15:43	
Air Blank		0.000	15:44	
Air Blank		0.000	15:45	
Subject Sample	# 2	0.270	15:46	
Breath Volume		2.558	15:46	
Air Blank		0.000	15:47	
Control Test		0.074	15:47	Control Outside Tolerance
Air Blank		0.000	15:47	

Data Download Date/Time: 04/26/2019 01:18

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENOVIL,GEDEON,
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BRENOVIL,GEDEON
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 14:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:19

SUBJECT'S LAST NAME: MOSHER

DATE/TIME OF TEST: 04/22/2019 16:04:12

Diagnostics Check		OK	15:54
Air Blank		0.000	15:55
Control Test		0.077	15:55
Air Blank		0.000	15:55
Subject Sample	# 1	0.276	15:56
Breath Volume		2.660	15:56
Air Blank		0.000	15:57
Air Blank		0.000	15:58
Subject Sample	# 2	0.240	15:59
Breath Volume		1.785	15:59
Air Blank		0.000	16:00
Air Blank		0.000	16:01
Subject Sample	# 3	0.273	16:02
Breath Volume		2.812	16:02
Air Blank		0.000	16:03
Control Test		0.077	16:03
Air Blank		0.000	16:04
Diagnostics Check		OK	16:04

Data Download Date/Time: 04/26/2019 01:18

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: MIGUEL

DATE/TIME OF TEST: 04/23/2019 02:59:55

Diagnostics Check		OK	02:53
Air Blank		0.000	02:53
Control Test		0.078	02:54
Air Blank		0.000	02:54
Subject Sample	# 1	0.137	02:55
Breath Volume		2.457	02:55
Air Blank		0.000	02:55
Air Blank		0.000	02:57
Subject Sample	# 2	0.122	02:58
Breath Volume		2.078	02:58
Air Blank		0.000	02:58
Control Test		0.077	02:59
Air Blank		0.000	02:59
Diagnostics Check		OK	02:59

Data Download Date/Time: 04/26/2019 01:18

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: BALL

DATE/TIME OF TEST: 04/27/2019 00:04:52

Diagnostics Check		OK	23:58
Air Blank		0.000	23:58
Control Test		0.077	23:59
Air Blank		0.000	23:59
Subject Sample	# 1	0.208	00:00
Breath Volume		2.042	00:00
Air Blank		0.000	00:00
Air Blank		0.000	00:02
Subject Sample	# 2	0.207	00:03
Breath Volume		3.714	00:03
Air Blank		0.000	00:03
Control Test		0.077	00:04
Air Blank		0.000	00:04
Diagnostics Check		OK	00:04

Data Download Date/Time: 05/27/2019 23:00

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FENTON,WILLIAM,B
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: FENTON,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 04/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: HOUGHTON

DATE/TIME OF TEST: 04/27/2019 23:16:43

Diagnostics Check		OK	23:10
Air Blank		0.000	23:11
Control Test		0.077	23:11
Air Blank		0.000	23:12
Subject Sample	# 1	0.000	23:12
Breath Volume		2.367	23:12
Air Blank		0.000	23:12
Air Blank		0.000	23:14
Subject Sample	# 2	0.000	23:15
Breath Volume		1.476	23:15
Air Blank		0.000	23:15
Control Test		0.078	23:16
Air Blank		0.000	23:16
Diagnostics Check		OK	23:16

Data Download Date/Time: 05/27/2019 23:00

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SHEINBERG,T
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 04/05/2019 01:59:16

Diagnostics Check		OK	01:52
Air Blank		0.000	01:53
Control Test		0.080	01:53
Air Blank		0.000	01:53
Subject Sample	# 1	0.000	01:54
Breath Volume		1.730	01:54
Air Blank		0.000	01:55
Air Blank		0.000	01:57
Subject Sample	# 2	0.000	01:57
Breath Volume		1.167	01:57
Air Blank		0.000	01:58
Control Test		0.081	01:58
Air Blank		0.000	01:59
Diagnostics Check		OK	01:59

Data Download Date/Time: 05/06/2019 14:26

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANDREWS,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:56

SUBJECT'S LAST NAME: LAZUKA

DATE/TIME OF TEST: 04/05/2019 20:22:56

Diagnostics Check		OK	20:16
Air Blank		0.000	20:17
Control Test		0.080	20:17
Air Blank		0.000	20:18
Subject Sample	# 1	0.008	20:18
Breath Volume		1.554	20:18
Air Blank		0.000	20:19
Air Blank		0.000	20:21
Subject Sample	# 2	0.007	20:21
Breath Volume		1.410	20:21
Air Blank		0.000	20:22
Control Test		0.080	20:22
Air Blank		0.000	20:22
Diagnostics Check		OK	20:22

Data Download Date/Time: 05/06/2019 14:26

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SULAIMAN,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: WEST

DATE/TIME OF TEST: 04/06/2019 02:32:37

Diagnostics Check		OK	02:25
Air Blank		0.000	02:25
Control Test		0.080	02:25
Air Blank		0.000	02:26
Subject Sample	# 1	0.174	02:28
Breath Volume		2.859	02:28
Air Blank		0.000	02:28
Air Blank		0.000	02:30
Subject Sample	# 2	0.164	02:31
Breath Volume		1.437	02:31
Air Blank		0.000	02:31
Control Test		0.080	02:32
Air Blank		0.000	02:32
Diagnostics Check		OK	02:32

Data Download Date/Time: 05/06/2019 14:26

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: BONILLA

DATE/TIME OF TEST: 04/06/2019 04:34:31

Diagnostics Check		OK	04:27
Air Blank		0.000	04:28
Control Test		0.080	04:28
Air Blank		0.000	04:29
Subject Sample	# 1	0.079	04:29
Breath Volume		1.816	04:29
Air Blank		0.000	04:30
Air Blank		0.000	04:32
Subject Sample	# 2	0.083	04:32
Breath Volume		2.632	04:32
Air Blank		0.000	04:33
Control Test		0.080	04:33
Air Blank		0.000	04:34
Diagnostics Check		OK	04:34

Data Download Date/Time: 05/06/2019 14:26

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ARCHAMBEAU,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: CRUZ CRUZ

DATE/TIME OF TEST: 04/07/2019 03:06:23

Diagnostics Check		OK	02:59
Air Blank		0.000	02:59
Control Test		0.080	02:59
Air Blank		0.000	03:00
Subject Sample	# 1	0.215	03:01
Breath Volume		2.164	03:01
Air Blank		0.000	03:02
Air Blank		0.000	03:04
Subject Sample	# 2	0.227	03:04
Breath Volume		1.820	03:04
Air Blank		0.000	03:05
Control Test		0.079	03:05
Air Blank		0.000	03:06
Diagnostics Check		OK	03:06

Data Download Date/Time: 05/06/2019 14:26

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES AVERY,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GIERMEK,T
ARRESTING OFFICER AGENCY: U of So FL Campus PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: HOFFMAN

DATE/TIME OF TEST: 04/07/2019 03:37:05

Diagnostics Check		OK	03:30
Air Blank		0.000	03:31
Control Test		0.080	03:31
Air Blank		0.000	03:32
Subject Sample	# 1	0.145	03:32
Breath Volume		2.390	03:32
Air Blank		0.000	03:33
Air Blank		0.000	03:35
Subject Sample	# 2	0.155	03:35
Breath Volume		2.730	03:35
Air Blank		0.000	03:36
Control Test		0.079	03:36
Air Blank		0.000	03:37
Diagnostics Check		OK	03:37

Data Download Date/Time: 05/06/2019 14:26

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: COSTA,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:03

SUBJECT'S LAST NAME: FRAUSTO-REYNA

DATE/TIME OF TEST: 04/07/2019 05:31:02

Diagnostics Check		OK	05:24
Air Blank		0.000	05:24
Control Test		0.080	05:25
Air Blank		0.000	05:25
Subject Sample	# 1	0.152	05:26
Breath Volume		1.617	05:26
Air Blank		0.000	05:27
Air Blank		0.000	05:28
Subject Sample	# 2	0.156	05:29
Breath Volume		2.027	05:29
Air Blank		0.000	05:30
Control Test		0.080	05:30
Air Blank		0.000	05:30
Diagnostics Check		OK	05:31

Data Download Date/Time: 05/06/2019 14:26

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SULAIMAN,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:51

SUBJECT'S LAST NAME: WEBB

DATE/TIME OF TEST: 04/07/2019 20:17:45

Diagnostics Check		OK	20:11
Air Blank		0.000	20:12
Control Test		0.080	20:12
Air Blank		0.000	20:12
Subject Sample	# 1	0.000	20:13
Breath Volume		1.437	20:13
Air Blank		0.000	20:13
Air Blank		0.000	20:15
Subject Sample	# 2	0.000	20:16
Breath Volume		1.914	20:16
Air Blank		0.000	20:16
Control Test		0.080	20:17
Air Blank		0.000	20:17
Diagnostics Check		OK	20:17

Data Download Date/Time: 05/06/2019 14:26

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES AVERY,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:04

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 04/07/2019 21:34:43

Diagnostics Check		OK	21:24
Air Blank		0.000	21:25
Control Test		0.080	21:25
Air Blank		0.000	21:26
Subject Sample	# 1	0.170	21:27
Breath Volume		1.402	21:27
Air Blank		0.000	21:28
Air Blank		0.000	21:30
Subject Sample	# 2	0.167	21:33
Breath Volume		1.328	21:33
Air Blank		0.000	21:33
Control Test		0.079	21:34
Air Blank		0.000	21:34
Diagnostics Check		OK	21:34

Data Download Date/Time: 05/06/2019 14:26

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES AVERY,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIMPKINS,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: MORALES PEREZ

DATE/TIME OF TEST: 04/08/2019 01:03:40

Diagnostics Check		OK	00:56
Air Blank		0.000	00:57
Control Test		0.079	00:57
Air Blank		0.000	00:57
Subject Sample	# 1	0.150	00:59
Breath Volume		2.660	00:59
Air Blank		0.000	00:59
Air Blank		0.000	01:01
Subject Sample	# 2	0.144	01:02
Breath Volume		1.730	01:02
Air Blank		0.000	01:02
Control Test		0.078	01:03
Air Blank		0.000	01:03
Diagnostics Check		OK	01:03

Data Download Date/Time: 05/06/2019 14:26

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 05:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:50

SUBJECT'S LAST NAME: ABDEL-RAHMAN

DATE/TIME OF TEST: 04/08/2019 06:17:50

Diagnostics Check		OK	06:11
Air Blank		0.000	06:11
Control Test		0.080	06:12
Air Blank		0.000	06:12
Subject Sample	# 1	0.111	06:13
Breath Volume		2.484	06:13
Air Blank		0.000	06:13
Air Blank		0.000	06:15
Subject Sample	# 2	0.107	06:16
Breath Volume		3.605	06:16
Air Blank		0.000	06:16
Control Test		0.079	06:17
Air Blank		0.000	06:17
Diagnostics Check		OK	06:17

Data Download Date/Time: 05/06/2019 14:26

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KLENOFSKY,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: WACH

DATE/TIME OF TEST: 04/09/2019 23:09:11

Diagnostics Check		OK	23:01
Air Blank		0.000	23:02
Control Test		0.080	23:02
Air Blank		0.000	23:02
Subject Sample	# 1	0.157	23:04
Breath Volume		1.453	23:04
Air Blank		0.000	23:05
Air Blank		0.000	23:07
Subject Sample	# 2	0.161	23:07
Breath Volume		2.242	23:07
Air Blank		0.000	23:08
Control Test		0.079	23:08
Air Blank		0.000	23:09
Diagnostics Check		OK	23:09

Data Download Date/Time: 05/06/2019 14:26

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HARRIS,S
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 04/10/2019 04:49:31

Diagnostics Check		OK	04:37	
Air Blank		0.000	04:38	
Control Test		0.077	04:38	
Air Blank		0.000	04:39	
Subject Sample	# 1	0.109	04:42	Volume Not Met
Breath Volume		1.046	04:42	
Air Blank		0.000	04:43	
Air Blank		0.000	04:44	
Subject Sample	# 2	0.114	04:47	Volume Not Met
Breath Volume		0.910	04:47	
Air Blank		0.000	04:48	
Control Test		0.076	04:48	
Air Blank		0.000	04:49	
Diagnostics Check		OK	04:49	

Data Download Date/Time: 05/06/2019 14:26

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ARCHAMBEAU,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:03

SUBJECT'S LAST NAME: CERROBLANCO CARDON

DATE/TIME OF TEST: 04/10/2019 19:31:19

Diagnostics Check		OK	19:24
Air Blank		0.000	19:24
Control Test		0.080	19:24
Air Blank		0.000	19:25
Subject Sample	# 1	0.222	19:26
Breath Volume		3.160	19:26
Air Blank		0.000	19:27
Air Blank		0.000	19:28
Subject Sample	# 2	0.228	19:29
Breath Volume		2.945	19:29
Air Blank		0.000	19:30
Control Test		0.079	19:30
Air Blank		0.000	19:31
Diagnostics Check		OK	19:31

Data Download Date/Time: 05/06/2019 14:26

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANDREWS,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: CHERAMIE

DATE/TIME OF TEST: 04/11/2019 02:21:01

Diagnostics Check		OK	02:11
Air Blank		0.000	02:11
Control Test		0.080	02:12
Air Blank		0.000	02:12
Subject Sample	# 1	0.175	02:14
Breath Volume		2.203	02:14
Air Blank		0.000	02:15
Air Blank		0.000	02:16
Subject Sample	# 2	0.166	02:19
Breath Volume		1.980	02:19
Air Blank		0.000	02:20
Control Test		0.079	02:20
Air Blank		0.000	02:20
Diagnostics Check		OK	02:21

Data Download Date/Time: 05/06/2019 14:26

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: CARDOSO GONCALVES

DATE/TIME OF TEST: 04/11/2019 04:13:18

Diagnostics Check		OK	04:09	
Air Blank		0.000	04:09	
Control Test		0.080	04:10	
Air Blank		0.000	04:10	
Subject Sample	# 1		04:11	Subject Test Refused
Breath Volume		0.000	04:11	
Air Blank		0.000	04:12	
Control Test		0.080	04:12	
Air Blank		0.000	04:13	
Diagnostics Check		OK	04:13	

Data Download Date/Time: 05/06/2019 14:27

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SCHAPPER,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:09

SUBJECT'S LAST NAME: VAUGHN

DATE/TIME OF TEST: 04/11/2019 18:37:20

Diagnostics Check		OK	18:31
Air Blank		0.000	18:31
Control Test		0.080	18:31
Air Blank		0.000	18:32
Subject Sample	# 1	0.183	18:32
Breath Volume		1.996	18:32
Air Blank		0.000	18:33
Air Blank		0.000	18:35
Subject Sample	# 2	0.170	18:35
Breath Volume		1.300	18:35
Air Blank		0.000	18:36
Control Test		0.080	18:36
Air Blank		0.000	18:37
Diagnostics Check		OK	18:37

Data Download Date/Time: 05/06/2019 14:27

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 04/12/2019 05:45:47

Diagnostics Check		OK	05:30	
Air Blank		0.000	05:31	
Control Test		0.081	05:31	
Air Blank		0.000	05:32	
Subject Sample	# 1	0.140	05:35	Volume Not Met
Breath Volume		0.531	05:35	
Air Blank		0.000	05:36	
Air Blank		0.000	05:37	
Subject Sample	# 2	0.140	05:39	
Breath Volume		1.355	05:39	
Air Blank		0.000	05:40	
Air Blank		0.000	05:41	
Subject Sample	# 3	0.146	05:44	
Breath Volume		1.429	05:44	
Air Blank		0.000	05:44	
Control Test		0.079	05:45	
Air Blank		0.000	05:45	
Diagnostics Check		OK	05:45	

Data Download Date/Time: 05/06/2019 14:27

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,REBECCA,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KROUSE,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 08:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:53

SUBJECT'S LAST NAME: CORSI

DATE/TIME OF TEST: 04/12/2019 09:25:14

Diagnostics Check		OK	09:15
Air Blank		0.000	09:16
Control Test		0.080	09:16
Air Blank		0.000	09:16
Subject Sample	# 1	0.081	09:19
Breath Volume		1.718	09:19
Air Blank		0.000	09:19
Air Blank		0.000	09:21
Subject Sample	# 2	0.082	09:23
Breath Volume		1.429	09:23
Air Blank		0.000	09:24
Control Test		0.080	09:24
Air Blank		0.000	09:25
Diagnostics Check		OK	09:25

Data Download Date/Time: 05/06/2019 14:27

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,TIMOTHY,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEE,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:10

SUBJECT'S LAST NAME: BARNHART

DATE/TIME OF TEST: 04/12/2019 16:39:35

Diagnostics Check		OK	16:31	
Air Blank		0.000	16:31	
Control Test		0.081	16:32	
Air Blank		0.000	16:32	
Subject Sample	# 1		16:35	No Sample Provided
Breath Volume		0.000	16:35	
Air Blank		0.000	16:36	
Air Blank			16:38	RFI Detect
Air Blank		0.000	16:38	
Control Test		0.081	16:38	
Air Blank		0.000	16:39	
Diagnostics Check		OK	16:39	

Data Download Date/Time: 05/06/2019 14:27

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,TIMOTHY,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEE,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:10

SUBJECT'S LAST NAME: BARNHART

DATE/TIME OF TEST: 04/12/2019 16:52:09

Diagnostics Check		OK	16:43	
Air Blank		0.000	16:43	
Control Test		0.080	16:44	
Air Blank		0.000	16:44	
Subject Sample	# 1	0.199	16:47	
Breath Volume		1.765	16:47	
Air Blank		0.000	16:47	
Air Blank		0.000	16:49	
Subject Sample	# 2		16:50	RFI Detect
Breath Volume		0.000	16:50	
Air Blank		0.000	16:51	
Control Test		0.079	16:51	
Air Blank		0.000	16:52	
Diagnostics Check		OK	16:52	

Data Download Date/Time: 05/06/2019 14:27

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 04/13/2019 02:35:31

Diagnostics Check		OK	02:27
Air Blank		0.000	02:27
Control Test		0.081	02:27
Air Blank		0.000	02:28
Subject Sample	# 1	0.197	02:29
Breath Volume		2.082	02:29
Air Blank		0.000	02:30
Air Blank		0.000	02:32
Subject Sample	# 2	0.191	02:33
Breath Volume		2.609	02:33
Air Blank		0.000	02:34
Control Test		0.080	02:34
Air Blank		0.000	02:35
Diagnostics Check		OK	02:35

Data Download Date/Time: 05/06/2019 14:27

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NAVE,R
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 04/13/2019 03:06:30

Diagnostics Check		OK	02:57	
Air Blank		0.000	02:58	
Control Test		0.080	02:58	
Air Blank		0.000	02:59	
Subject Sample	# 1		03:02	No Sample Provided
Breath Volume		0.000	03:02	
Air Blank		0.000	03:02	
Air Blank		0.000	03:04	
Subject Sample	# 2		03:05	Subject Test Refused
Breath Volume		0.000	03:05	
Air Blank		0.000	03:05	
Control Test		0.081	03:05	
Air Blank		0.000	03:06	
Diagnostics Check		OK	03:06	

Data Download Date/Time: 05/06/2019 14:27

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WOLSTENHOLM,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: BARI

DATE/TIME OF TEST: 04/13/2019 04:28:47

Diagnostics Check		OK	04:22
Air Blank		0.000	04:22
Control Test		0.081	04:22
Air Blank		0.000	04:23
Subject Sample	# 1	0.171	04:24
Breath Volume		1.246	04:24
Air Blank		0.000	04:24
Air Blank		0.000	04:26
Subject Sample	# 2	0.176	04:27
Breath Volume		1.375	04:27
Air Blank		0.000	04:27
Control Test		0.080	04:28
Air Blank		0.000	04:28
Diagnostics Check		OK	04:28

Data Download Date/Time: 05/06/2019 14:27

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FISHER,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:36

SUBJECT'S LAST NAME: CASTILLO GOMEZ

DATE/TIME OF TEST: 04/13/2019 05:03:19

Diagnostics Check		OK	04:56
Air Blank		0.000	04:57
Control Test		0.080	04:57
Air Blank		0.000	04:58
Subject Sample	# 1	0.152	04:58
Breath Volume		2.347	04:58
Air Blank		0.000	04:59
Air Blank		0.000	05:01
Subject Sample	# 2	0.150	05:01
Breath Volume		1.820	05:01
Air Blank		0.000	05:02
Control Test		0.079	05:02
Air Blank		0.000	05:03
Diagnostics Check		OK	05:03

Data Download Date/Time: 05/06/2019 14:27

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CAMPBELL,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: MADISON

DATE/TIME OF TEST: 04/14/2019 02:51:39

Diagnostics Check		OK	02:44
Air Blank		0.000	02:45
Control Test		0.081	02:45
Air Blank		0.000	02:45
Subject Sample	# 1	0.050	02:47
Breath Volume		1.894	02:47
Air Blank		0.000	02:47
Air Blank		0.000	02:49
Subject Sample	# 2	0.048	02:50
Breath Volume		2.277	02:50
Air Blank		0.000	02:50
Control Test		0.080	02:51
Air Blank		0.000	02:51
Diagnostics Check		OK	02:51

Data Download Date/Time: 05/06/2019 14:27

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: WAGNER

DATE/TIME OF TEST: 04/14/2019 04:18:53

Diagnostics Check		OK	04:12
Air Blank		0.000	04:13
Control Test		0.081	04:13
Air Blank		0.000	04:14
Subject Sample	# 1	0.160	04:14
Breath Volume		1.417	04:14
Air Blank		0.000	04:15
Air Blank		0.000	04:16
Subject Sample	# 2	0.159	04:17
Breath Volume		1.449	04:17
Air Blank		0.000	04:17
Control Test		0.081	04:18
Air Blank		0.000	04:18
Diagnostics Check		OK	04:18

Data Download Date/Time: 05/06/2019 14:27

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HOBSON,D
ARRESTING OFFICER AGENCY: U of So FL Campus PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: GRAVES

DATE/TIME OF TEST: 04/14/2019 04:56:34

Diagnostics Check		OK	04:50
Air Blank		0.000	04:50
Control Test		0.080	04:50
Air Blank		0.000	04:51
Subject Sample	# 1	0.151	04:52
Breath Volume		3.078	04:52
Air Blank		0.000	04:52
Air Blank		0.000	04:54
Subject Sample	# 2	0.150	04:54
Breath Volume		3.167	04:54
Air Blank		0.000	04:55
Control Test		0.080	04:55
Air Blank		0.000	04:56
Diagnostics Check		OK	04:56

Data Download Date/Time: 05/06/2019 14:27

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,ANGELA,V
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MACIAS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 13:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:44

SUBJECT'S LAST NAME: MULLINS

DATE/TIME OF TEST: 04/14/2019 14:12:30

Diagnostics Check		OK	14:04
Air Blank		0.000	14:05
Control Test		0.081	14:05
Air Blank		0.000	14:06
Subject Sample	# 1	0.122	14:07
Breath Volume		1.832	14:07
Air Blank		0.000	14:07
Air Blank		0.000	14:09
Subject Sample	# 2	0.124	14:10
Breath Volume		1.941	14:10
Air Blank		0.000	14:11
Control Test		0.080	14:11
Air Blank		0.000	14:12
Diagnostics Check		OK	14:12

Data Download Date/Time: 05/06/2019 14:27

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: CONKLIN

DATE/TIME OF TEST: 04/14/2019 23:21:40

Diagnostics Check		OK	23:13
Air Blank		0.000	23:13
Control Test		0.081	23:13
Air Blank		0.000	23:14
Subject Sample	# 1	0.075	23:16
Breath Volume		2.730	23:16
Air Blank		0.000	23:17
Air Blank		0.000	23:19
Subject Sample	# 2	0.071	23:20
Breath Volume		3.046	23:20
Air Blank		0.000	23:20
Control Test		0.080	23:21
Air Blank		0.000	23:21
Diagnostics Check		OK	23:21

Data Download Date/Time: 05/06/2019 14:27

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKIPPER,ROGER,G
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: REICHERT,C 4703
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 07:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:08

SUBJECT'S LAST NAME: BURNETT

DATE/TIME OF TEST: 04/15/2019 08:37:50

Diagnostics Check		OK	08:30
Air Blank		0.000	08:31
Control Test		0.080	08:31
Air Blank		0.000	08:32
Subject Sample	# 1	0.000	08:33
Breath Volume		2.195	08:33
Air Blank		0.000	08:33
Air Blank		0.000	08:35
Subject Sample	# 2	0.000	08:36
Breath Volume		2.445	08:36
Air Blank		0.000	08:36
Control Test		0.080	08:37
Air Blank		0.000	08:37
Diagnostics Check		OK	08:37

Data Download Date/Time: 05/06/2019 14:27

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SULAIMAN,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: WYMAN

DATE/TIME OF TEST: 04/16/2019 00:34:00

Diagnostics Check	OK	00:32	
Air Blank	0.000	00:33	
Control Test	0.080	00:33	
Air Blank		00:33	RFI Detect
Air Blank	0.000	00:34	

Data Download Date/Time: 05/06/2019 14:27

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SULAIMAN,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: WYMAN

DATE/TIME OF TEST: 04/16/2019 00:41:39

Diagnostics Check		OK	00:35
Air Blank		0.000	00:35
Control Test		0.080	00:35
Air Blank		0.000	00:36
Subject Sample	# 1	0.133	00:37
Breath Volume		1.347	00:37
Air Blank		0.000	00:37
Air Blank		0.000	00:39
Subject Sample	# 2	0.138	00:40
Breath Volume		1.894	00:40
Air Blank		0.000	00:40
Control Test		0.080	00:41
Air Blank		0.000	00:41
Diagnostics Check		OK	00:41

Data Download Date/Time: 05/06/2019 14:27

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARRELL,Q
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: TILLMAN

DATE/TIME OF TEST: 04/16/2019 01:26:56

Diagnostics Check		OK	01:19
Air Blank		0.000	01:19
Control Test		0.080	01:20
Air Blank		0.000	01:20
Subject Sample	# 1	0.193	01:21
Breath Volume		2.636	01:21
Air Blank		0.000	01:22
Air Blank		0.000	01:24
Subject Sample	# 2	0.203	01:25
Breath Volume		3.183	01:25
Air Blank		0.000	01:26
Control Test		0.080	01:26
Air Blank		0.000	01:26
Diagnostics Check		OK	01:26

Data Download Date/Time: 05/06/2019 14:27

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LYON,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 04/19/2019 01:02:22

Diagnostics Check		OK	00:53
Air Blank		0.000	00:54
Control Test		0.081	00:54
Air Blank		0.000	00:55
Subject Sample	# 1	0.200	00:56
Breath Volume		1.160	00:56
Air Blank		0.000	00:57
Air Blank		0.000	00:59
Subject Sample	# 2	0.200	01:00
Breath Volume		2.656	01:00
Air Blank		0.000	01:01
Control Test		0.080	01:01
Air Blank		0.000	01:02
Diagnostics Check		OK	01:02

Data Download Date/Time: 05/06/2019 14:27

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KROUSE,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: PASQUALE

DATE/TIME OF TEST: 04/19/2019 23:25:32

Diagnostics Check		OK	23:18
Air Blank		0.000	23:18
Control Test		0.081	23:19
Air Blank		0.000	23:19
Subject Sample	# 1	0.181	23:20
Breath Volume		2.390	23:20
Air Blank		0.000	23:21
Air Blank		0.000	23:23
Subject Sample	# 2	0.165	23:23
Breath Volume		1.398	23:23
Air Blank		0.000	23:24
Control Test		0.080	23:24
Air Blank		0.000	23:25
Diagnostics Check		OK	23:25

Data Download Date/Time: 05/06/2019 14:27

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALL,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: GONZALEZ LOPEZ

DATE/TIME OF TEST: 04/20/2019 00:54:21

Diagnostics Check		OK	00:47
Air Blank		0.000	00:48
Control Test		0.081	00:48
Air Blank		0.000	00:49
Subject Sample	# 1	0.159	00:49
Breath Volume		1.144	00:49
Air Blank		0.000	00:50
Air Blank		0.000	00:52
Subject Sample	# 2	0.155	00:52
Breath Volume		1.281	00:52
Air Blank		0.000	00:53
Control Test		0.080	00:53
Air Blank		0.000	00:54
Diagnostics Check		OK	00:54

Data Download Date/Time: 05/06/2019 14:27

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DRISCOLL,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 04/20/2019 04:47:18

Diagnostics Check		OK	04:40
Air Blank		0.000	04:41
Control Test		0.081	04:41
Air Blank		0.000	04:42
Subject Sample	# 1	0.053	04:42
Breath Volume		1.328	04:42
Air Blank		0.000	04:43
Air Blank		0.000	04:45
Subject Sample	# 2	0.056	04:45
Breath Volume		2.976	04:45
Air Blank		0.000	04:46
Control Test		0.081	04:46
Air Blank		0.000	04:47
Diagnostics Check		OK	04:47

Data Download Date/Time: 05/06/2019 14:27

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANDREWS,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 04/20/2019 23:27:18

Diagnostics Check		OK	23:20
Air Blank		0.000	23:21
Control Test		0.080	23:21
Air Blank		0.000	23:22
Subject Sample	# 1	0.142	23:22
Breath Volume		1.308	23:22
Air Blank		0.000	23:23
Air Blank		0.000	23:25
Subject Sample	# 2	0.134	23:25
Breath Volume		1.433	23:25
Air Blank		0.000	23:26
Control Test		0.079	23:26
Air Blank		0.000	23:27
Diagnostics Check		OK	23:27

Data Download Date/Time: 05/06/2019 14:27

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES AVERY,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALL,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: ORANGE JR

DATE/TIME OF TEST: 04/21/2019 02:23:20

Diagnostics Check		OK	02:17
Air Blank		0.000	02:17
Control Test		0.081	02:18
Air Blank		0.000	02:18
Subject Sample	# 1	0.019	02:18
Breath Volume		2.496	02:18
Air Blank		0.000	02:19
Air Blank		0.000	02:21
Subject Sample	# 2	0.019	02:21
Breath Volume		3.289	02:21
Air Blank		0.000	02:22
Control Test		0.080	02:22
Air Blank		0.000	02:23
Diagnostics Check		OK	02:23

Data Download Date/Time: 05/06/2019 14:27

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES AVERY,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DOBSON,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: MCKINNON

DATE/TIME OF TEST: 04/21/2019 02:38:02

Diagnostics Check		OK	02:32
Air Blank		0.000	02:32
Control Test		0.080	02:32
Air Blank		0.000	02:33
Subject Sample	# 1	0.092	02:33
Breath Volume		1.316	02:33
Air Blank		0.000	02:34
Air Blank		0.000	02:36
Subject Sample	# 2	0.092	02:36
Breath Volume		1.539	02:36
Air Blank		0.000	02:37
Control Test		0.080	02:37
Air Blank		0.000	02:37
Diagnostics Check		OK	02:38

Data Download Date/Time: 05/06/2019 14:27

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LARA,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 04/21/2019 04:20:11

Diagnostics Check		OK	04:13
Air Blank		0.000	04:14
Control Test		0.080	04:14
Air Blank		0.000	04:14
Subject Sample	# 1	0.156	04:15
Breath Volume		3.804	04:15
Air Blank		0.000	04:16
Air Blank		0.000	04:18
Subject Sample	# 2	0.145	04:18
Breath Volume		2.464	04:18
Air Blank		0.000	04:19
Control Test		0.079	04:19
Air Blank		0.000	04:20
Diagnostics Check		OK	04:20

Data Download Date/Time: 05/06/2019 14:27

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: MULLINS III

DATE/TIME OF TEST: 04/22/2019 00:50:14

Diagnostics Check		OK	00:43
Air Blank		0.000	00:44
Control Test		0.081	00:44
Air Blank		0.000	00:44
Subject Sample	# 1	0.204	00:45
Breath Volume		3.292	00:45
Air Blank		0.000	00:46
Air Blank		0.000	00:48
Subject Sample	# 2	0.199	00:48
Breath Volume		3.484	00:48
Air Blank		0.000	00:49
Control Test		0.080	00:49
Air Blank		0.000	00:50
Diagnostics Check		OK	00:50

Data Download Date/Time: 05/06/2019 14:27

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: SWANSON

DATE/TIME OF TEST: 04/24/2019 01:48:43

Diagnostics Check		OK	01:34	
Air Blank		0.000	01:34	
Control Test		0.081	01:34	
Air Blank		0.000	01:35	
Subject Sample	# 1	0.164	01:37	
Breath Volume		1.273	01:37	
Air Blank		0.000	01:37	
Air Blank		0.000	01:39	
Subject Sample	# 2	0.120	01:42	Volume Not Met
Breath Volume		0.394	01:42	
Air Blank		0.000	01:43	
Air Blank		0.000	01:45	
Subject Sample	# 3	0.156	01:47	
Breath Volume		1.421	01:47	
Air Blank		0.000	01:47	
Control Test		0.078	01:48	
Air Blank		0.000	01:48	
Diagnostics Check		OK	01:48	

Data Download Date/Time: 05/06/2019 14:27

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: DEL ANGEL HERNANDEZ

DATE/TIME OF TEST: 04/24/2019 22:54:19

Diagnostics Check		OK	22:46
Air Blank		0.000	22:46
Control Test		0.080	22:46
Air Blank		0.000	22:47
Subject Sample	# 1	0.158	22:49
Breath Volume		1.554	22:49
Air Blank		0.000	22:49
Air Blank		0.000	22:51
Subject Sample	# 2	0.159	22:52
Breath Volume		1.394	22:52
Air Blank		0.000	22:53
Control Test		0.079	22:53
Air Blank		0.000	22:54
Diagnostics Check		OK	22:54

Data Download Date/Time: 05/06/2019 14:27

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: WILKERSON

DATE/TIME OF TEST: 04/25/2019 00:55:13

Diagnostics Check		OK	00:47
Air Blank		0.000	00:48
Control Test		0.080	00:48
Air Blank		0.000	00:48
Subject Sample	# 1	0.092	00:50
Breath Volume		2.058	00:50
Air Blank		0.000	00:51
Air Blank		0.000	00:53
Subject Sample	# 2	0.089	00:53
Breath Volume		2.195	00:53
Air Blank		0.000	00:54
Control Test		0.080	00:54
Air Blank		0.000	00:55
Diagnostics Check		OK	00:55

Data Download Date/Time: 05/06/2019 14:27

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 04/25/2019 03:31:49

Diagnostics Check		OK	03:21
Air Blank		0.000	03:21
Control Test		0.081	03:21
Air Blank		0.000	03:22
Subject Sample	# 1	0.261	03:23
Breath Volume		1.632	03:23
Air Blank		0.000	03:24
Air Blank		0.000	03:25
Subject Sample	# 2	0.239	03:26
Breath Volume		2.300	03:26
Air Blank		0.000	03:27
Air Blank		0.000	03:29
Subject Sample	# 3	0.237	03:30
Breath Volume		1.242	03:30
Air Blank		0.000	03:30
Control Test		0.079	03:31
Air Blank		0.000	03:31
Diagnostics Check		OK	03:31

Data Download Date/Time: 05/06/2019 14:27

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VANTREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: PIEMME

DATE/TIME OF TEST: 04/26/2019 00:59:54

Diagnostics Check		OK	00:53
Air Blank		0.000	00:53
Control Test		0.081	00:54
Air Blank		0.000	00:54
Subject Sample	# 1	0.173	00:55
Breath Volume		2.570	00:55
Air Blank		0.000	00:56
Air Blank		0.000	00:57
Subject Sample	# 2	0.165	00:58
Breath Volume		2.164	00:58
Air Blank		0.000	00:58
Control Test		0.080	00:59
Air Blank		0.000	00:59
Diagnostics Check		OK	00:59

Data Download Date/Time: 05/06/2019 14:27

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKIPPER,ROGER,G
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DRISCOLL,B 251509
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 05:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:29

SUBJECT'S LAST NAME: ROMERO-OLIVERAS

DATE/TIME OF TEST: 04/26/2019 06:57:02

Diagnostics Check		OK	06:50
Air Blank		0.000	06:50
Control Test		0.080	06:51
Air Blank		0.000	06:51
Subject Sample	# 1	0.194	06:52
Breath Volume		2.046	06:52
Air Blank		0.000	06:53
Air Blank		0.000	06:54
Subject Sample	# 2	0.182	06:55
Breath Volume		1.222	06:55
Air Blank		0.000	06:56
Control Test		0.079	06:56
Air Blank		0.000	06:56
Diagnostics Check		OK	06:57

Data Download Date/Time: 05/06/2019 14:27

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,REBECCA,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 06:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:34

SUBJECT'S LAST NAME: ELLINGTON

DATE/TIME OF TEST: 04/26/2019 08:01:33

Diagnostics Check		OK	07:55
Air Blank		0.000	07:55
Control Test		0.080	07:55
Air Blank		0.000	07:56
Subject Sample	# 1	0.000	07:57
Breath Volume		1.750	07:57
Air Blank		0.000	07:57
Air Blank		0.000	07:59
Subject Sample	# 2	0.000	08:00
Breath Volume		2.261	08:00
Air Blank		0.000	08:00
Control Test		0.080	08:00
Air Blank		0.000	08:01
Diagnostics Check		OK	08:01

Data Download Date/Time: 05/06/2019 14:27

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: GARCIA-PEREZ

DATE/TIME OF TEST: 04/26/2019 23:01:26

Diagnostics Check		OK	22:53
Air Blank		0.000	22:54
Control Test		0.081	22:54
Air Blank		0.000	22:54
Subject Sample	# 1	0.206	22:55
Breath Volume		1.285	22:55
Air Blank		0.000	22:56
Air Blank		0.000	22:57
Subject Sample	# 2	0.198	22:59
Breath Volume		1.492	22:59
Air Blank		0.000	23:00
Control Test		0.080	23:00
Air Blank		0.000	23:01
Diagnostics Check		OK	23:01

Data Download Date/Time: 05/06/2019 14:27

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KLENOFSKY,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: FERIES

DATE/TIME OF TEST: 04/27/2019 02:55:32

Diagnostics Check		OK	02:49
Air Blank		0.000	02:49
Control Test		0.081	02:50
Air Blank		0.000	02:50
Subject Sample	# 1	0.143	02:51
Breath Volume		2.949	02:51
Air Blank		0.000	02:51
Air Blank		0.000	02:53
Subject Sample	# 2	0.141	02:53
Breath Volume		3.046	02:53
Air Blank		0.000	02:54
Control Test		0.080	02:54
Air Blank		0.000	02:55
Diagnostics Check		OK	02:55

Data Download Date/Time: 05/06/2019 14:27

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: CORZO

DATE/TIME OF TEST: 04/28/2019 01:54:28

Diagnostics Check		OK	01:48
Air Blank		0.000	01:48
Control Test		0.081	01:48
Air Blank		0.000	01:49
Subject Sample	# 1	0.179	01:49
Breath Volume		1.664	01:49
Air Blank		0.000	01:50
Air Blank		0.000	01:52
Subject Sample	# 2	0.181	01:52
Breath Volume		2.242	01:52
Air Blank		0.000	01:53
Control Test		0.080	01:53
Air Blank		0.000	01:54
Diagnostics Check		OK	01:54

Data Download Date/Time: 05/06/2019 14:28

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CAMPBELL,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: REYES RODRIGUEZ

DATE/TIME OF TEST: 04/28/2019 03:51:13

Diagnostics Check		OK	03:43
Air Blank		0.000	03:44
Control Test		0.080	03:44
Air Blank		0.000	03:45
Subject Sample	# 1	0.183	03:46
Breath Volume		3.015	03:46
Air Blank		0.000	03:47
Air Blank		0.000	03:49
Subject Sample	# 2	0.187	03:49
Breath Volume		1.476	03:49
Air Blank		0.000	03:50
Control Test		0.079	03:50
Air Blank		0.000	03:51
Diagnostics Check		OK	03:51

Data Download Date/Time: 05/06/2019 14:28

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: BASORA HERNANDEZ

DATE/TIME OF TEST: 04/28/2019 04:25:07

Diagnostics Check		OK	04:18
Air Blank		0.000	04:19
Control Test		0.079	04:19
Air Blank		0.000	04:20
Subject Sample	# 1	0.000	04:20
Breath Volume		2.976	04:20
Air Blank		0.000	04:21
Air Blank		0.000	04:23
Subject Sample	# 2	0.000	04:23
Breath Volume		3.437	04:23
Air Blank		0.000	04:24
Control Test		0.078	04:24
Air Blank		0.000	04:25
Diagnostics Check		OK	04:25

Data Download Date/Time: 05/06/2019 14:28

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LYON,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: MONGIOVI

DATE/TIME OF TEST: 04/28/2019 04:42:34

Diagnostics Check		OK	04:35
Air Blank		0.000	04:35
Control Test		0.080	04:36
Air Blank		0.000	04:36
Subject Sample	# 1	0.161	04:38
Breath Volume		2.566	04:38
Air Blank		0.000	04:38
Air Blank		0.000	04:40
Subject Sample	# 2	0.156	04:40
Breath Volume		1.906	04:40
Air Blank		0.000	04:41
Control Test		0.079	04:41
Air Blank		0.000	04:42
Diagnostics Check		OK	04:42

Data Download Date/Time: 05/06/2019 14:28

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DE JESUS,V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: HUBBARD

DATE/TIME OF TEST: 04/28/2019 05:15:22

Diagnostics Check		OK	05:08
Air Blank		0.000	05:09
Control Test		0.079	05:09
Air Blank		0.000	05:09
Subject Sample	# 1	0.112	05:10
Breath Volume		2.355	05:10
Air Blank		0.000	05:11
Air Blank		0.000	05:13
Subject Sample	# 2	0.113	05:13
Breath Volume		2.601	05:13
Air Blank		0.000	05:14
Control Test		0.080	05:14
Air Blank		0.000	05:15
Diagnostics Check		OK	05:15

Data Download Date/Time: 05/06/2019 14:28

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: THORNE,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:01

SUBJECT'S LAST NAME: DANIELS

DATE/TIME OF TEST: 04/28/2019 20:38:27

Diagnostics Check		OK	20:22	
Air Blank		0.000	20:22	
Control Test		0.081	20:22	
Air Blank		0.000	20:23	
Subject Sample	# 1	0.129	20:26	Volume Not Met
Breath Volume		0.191	20:26	
Air Blank		0.000	20:27	
Air Blank		0.000	20:28	
Subject Sample	# 2	0.133	20:31	
Breath Volume		1.593	20:31	
Air Blank		0.000	20:32	
Air Blank		0.000	20:34	
Subject Sample	# 3	0.123	20:36	
Breath Volume		1.648	20:36	
Air Blank		0.000	20:37	
Control Test		0.078	20:37	
Air Blank		0.000	20:38	
Diagnostics Check		OK	20:38	

Data Download Date/Time: 05/06/2019 14:28

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 04/30/2019 00:52:29

Diagnostics Check		OK	00:46
Air Blank		0.000	00:46
Control Test		0.081	00:46
Air Blank		0.000	00:47
Subject Sample	# 1	0.149	00:47
Breath Volume		1.925	00:47
Air Blank		0.000	00:48
Air Blank		0.000	00:50
Subject Sample	# 2	0.147	00:50
Breath Volume		1.664	00:50
Air Blank		0.000	00:51
Control Test		0.080	00:51
Air Blank		0.000	00:52
Diagnostics Check		OK	00:52

Data Download Date/Time: 05/06/2019 14:28

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANDREWS,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: ENG

DATE/TIME OF TEST: 04/30/2019 03:54:15

Diagnostics Check		OK	03:47
Air Blank		0.000	03:47
Control Test		0.080	03:48
Air Blank		0.000	03:48
Subject Sample	# 1	0.248	03:49
Breath Volume		2.917	03:49
Air Blank		0.000	03:50
Air Blank		0.000	03:52
Subject Sample	# 2	0.265	03:52
Breath Volume		3.734	03:52
Air Blank		0.000	03:53
Control Test		0.080	03:53
Air Blank		0.000	03:54
Diagnostics Check		OK	03:54

Data Download Date/Time: 05/06/2019 14:28

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRUSE,KRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: REBANDO,BRIAN
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: BONO

DATE/TIME OF TEST: 04/02/2019 02:02:52

Diagnostics Check		OK	01:54
Air Blank		0.000	01:54
Control Test		0.081	01:54
Air Blank		0.000	01:55
Subject Sample	# 1	0.050	01:57
Breath Volume		3.429	01:57
Air Blank		0.000	01:58
Air Blank		0.000	02:00
Subject Sample	# 2	0.049	02:01
Breath Volume		2.449	02:01
Air Blank		0.000	02:01
Control Test		0.081	02:02
Air Blank		0.000	02:02
Diagnostics Check		OK	02:02

Data Download Date/Time: 04/10/2019 23:33

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALHOUN,JOHN,A
BREATH TEST OPERATOR AGENCY: Astatula PD
ARRESTING OFFICER: KING,SHADRACH
ARRESTING OFFICER AGENCY: Mascotte PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: HOGAN

DATE/TIME OF TEST: 04/02/2019 03:41:56

Diagnostics Check		OK	03:37	
Air Blank		0.000	03:37	
Control Test		0.080	03:37	
Air Blank		0.000	03:38	
Subject Sample	# 1		03:40	Subject Test Refused
Breath Volume		0.000	03:40	
Air Blank		0.000	03:41	
Control Test		0.080	03:41	
Air Blank		0.000	03:41	
Diagnostics Check		OK	03:41	

Data Download Date/Time: 04/10/2019 23:33

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALHOUN,JOHN,A
BREATH TEST OPERATOR AGENCY: Astatula PD
ARRESTING OFFICER: ROGERS,MICHELLE
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: NUNEZ

DATE/TIME OF TEST: 04/02/2019 04:19:58

Diagnostics Check		OK	04:13
Air Blank		0.000	04:13
Control Test		0.081	04:14
Air Blank		0.000	04:14
Subject Sample	# 1	0.164	04:15
Breath Volume		2.464	04:15
Air Blank		0.000	04:15
Air Blank		0.000	04:17
Subject Sample	# 2	0.161	04:18
Breath Volume		3.824	04:18
Air Blank		0.000	04:19
Control Test		0.081	04:19
Air Blank		0.000	04:19
Diagnostics Check		OK	04:19

Data Download Date/Time: 04/10/2019 23:33

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITCHER,DARRYL,P
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: PITCHER,DARRYL
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: MANN

DATE/TIME OF TEST: 04/06/2019 02:24:39

Diagnostics Check		OK	02:21	
Air Blank		0.000	02:21	
Control Test		0.081	02:22	
Air Blank		0.000	02:22	
Subject Sample	# 1		02:23	Subject Test Refused
Breath Volume		0.000	02:23	
Air Blank		0.000	02:23	
Control Test		0.081	02:24	
Air Blank		0.000	02:24	
Diagnostics Check		OK	02:24	

Data Download Date/Time: 04/10/2019 23:33

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITCHER,DARRYL,
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: APONTE,JOSH
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: WALLACE

DATE/TIME OF TEST: 04/06/2019 03:10:12

Diagnostics Check		OK	03:03
Air Blank		0.000	03:04
Control Test		0.080	03:04
Air Blank		0.000	03:05
Subject Sample	# 1	0.111	03:05
Breath Volume		2.585	03:05
Air Blank		0.000	03:06
Air Blank		0.000	03:08
Subject Sample	# 2	0.108	03:08
Breath Volume		2.437	03:08
Air Blank		0.000	03:09
Control Test		0.081	03:09
Air Blank		0.000	03:10
Diagnostics Check		OK	03:10

Data Download Date/Time: 04/10/2019 23:33

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRUSE,KRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: REBANDO,BRIAN
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: SANTANA

DATE/TIME OF TEST: 04/06/2019 04:13:40

Diagnostics Check	OK	04:12	
Air Blank	0.000	04:12	
Control Test	0.079	04:13	Interferent Detect
Air Blank	0.000	04:13	

Data Download Date/Time: 04/10/2019 23:33

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRUSE,KRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: REBANDO,BRIAN
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: SANTANA

DATE/TIME OF TEST: 04/06/2019 04:20:05

Diagnostics Check	OK	04:18	
Air Blank	0.000	04:19	
Control Test	0.083	04:19	Interferent Detect
Air Blank	0.000	04:20	

Data Download Date/Time: 04/10/2019 23:33

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRUSE,KRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: KRUSE,KRISTOPHER
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 04/06/2019 04:30:09

Diagnostics Check		OK	04:24	
Air Blank		0.000	04:25	
Control Test		0.080	04:25	
Air Blank		0.000	04:26	
Subject Sample	# 1		04:28	Subject Test Refused
Breath Volume		0.000	04:28	
Air Blank		0.000	04:29	
Control Test		0.080	04:29	
Air Blank		0.000	04:30	
Diagnostics Check		OK	04:30	

Data Download Date/Time: 04/10/2019 23:33

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRUSE,KRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: REBANDO,BRIAN
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: SANTANA

DATE/TIME OF TEST: 04/06/2019 04:37:52

Diagnostics Check		OK	04:34	
Air Blank		0.000	04:35	
Control Test		0.080	04:35	
Air Blank		0.000	04:36	
Subject Sample	# 1		04:36	Subject Test Refused
Breath Volume		0.000	04:36	
Air Blank		0.000	04:36	
Control Test		0.080	04:37	
Air Blank		0.000	04:37	
Diagnostics Check		OK	04:37	

Data Download Date/Time: 04/10/2019 23:33

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALHOUN,JOHN,A
BREATH TEST OPERATOR AGENCY: Astatula PD
ARRESTING OFFICER: PRINGLE,AUSTIN
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 18:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:42

SUBJECT'S LAST NAME: MONTERO

DATE/TIME OF TEST: 04/06/2019 20:10:19

Diagnostics Check		OK	20:02
Air Blank		0.000	20:03
Control Test		0.080	20:03
Air Blank		0.000	20:04
Subject Sample	# 1	0.105	20:05
Breath Volume		2.527	20:05
Air Blank		0.000	20:06
Air Blank		0.000	20:08
Subject Sample	# 2	0.109	20:08
Breath Volume		1.457	20:08
Air Blank		0.000	20:09
Control Test		0.081	20:09
Air Blank		0.000	20:10
Diagnostics Check		OK	20:10

Data Download Date/Time: 04/10/2019 23:33

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORMAN,GERALD,E
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: PARRISH,C
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: LERMA

DATE/TIME OF TEST: 04/07/2019 03:04:47

Diagnostics Check		OK	02:56
Air Blank		0.000	02:57
Control Test		0.080	02:57
Air Blank		0.000	02:57
Subject Sample	# 1	0.097	02:58
Breath Volume		1.273	02:58
Air Blank		0.000	02:59
Air Blank		0.000	03:01
Subject Sample	# 2	0.102	03:03
Breath Volume		1.414	03:03
Air Blank		0.000	03:03
Control Test		0.084	03:04
Air Blank		0.000	03:04
Diagnostics Check		OK	03:04

Data Download Date/Time: 04/10/2019 23:33

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORMAN,CORMAN,E
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CARROLL,F
ARRESTING OFFICER AGENCY: Fruitland Park PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: FOSTER

DATE/TIME OF TEST: 04/07/2019 03:23:49

Diagnostics Check		OK	03:20	
Air Blank		0.000	03:21	
Control Test		0.080	03:21	
Air Blank		0.000	03:22	
Subject Sample	# 1		03:22	Subject Test Refused
Breath Volume		0.000	03:22	
Air Blank		0.000	03:22	
Control Test		0.080	03:23	
Air Blank		0.000	03:23	
Diagnostics Check		OK	03:23	

Data Download Date/Time: 04/10/2019 23:33

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORMAN,GERALD,E
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: APONTE,J
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: GUELL

DATE/TIME OF TEST: 04/07/2019 03:32:56

Diagnostics Check		OK	03:29	
Air Blank		0.000	03:30	
Control Test		0.080	03:30	
Air Blank		0.000	03:31	
Subject Sample	# 1		03:31	Subject Test Refused
Breath Volume		0.000	03:31	
Air Blank		0.000	03:32	
Control Test		0.080	03:32	
Air Blank		0.000	03:32	
Diagnostics Check		OK	03:32	

Data Download Date/Time: 04/10/2019 23:33

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALHOUN,JOHN,A
BREATH TEST OPERATOR AGENCY: Astatula PD
ARRESTING OFFICER: CALHOUN,JOHN
ARRESTING OFFICER AGENCY: Astatula PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: SHOFFNER

DATE/TIME OF TEST: 04/08/2019 05:47:15

Diagnostics Check		OK	05:40
Air Blank		0.000	05:41
Control Test		0.080	05:41
Air Blank		0.000	05:42
Subject Sample	# 1	0.158	05:42
Breath Volume		1.828	05:42
Air Blank		0.000	05:43
Air Blank		0.000	05:45
Subject Sample	# 2	0.158	05:45
Breath Volume		1.511	05:45
Air Blank		0.000	05:46
Control Test		0.080	05:46
Air Blank		0.000	05:47
Diagnostics Check		OK	05:47

Data Download Date/Time: 04/10/2019 23:33

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALHOUN,JOHN,A
BREATH TEST OPERATOR AGENCY: Astatula PD
ARRESTING OFFICER: CLARK,KURTIS
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: BALATOVIS

DATE/TIME OF TEST: 04/10/2019 23:49:42

Diagnostics Check		OK	23:43
Air Blank		0.000	23:43
Control Test		0.080	23:43
Air Blank		0.000	23:44
Subject Sample	# 1	0.214	23:45
Breath Volume		3.945	23:45
Air Blank		0.000	23:45
Air Blank		0.000	23:47
Subject Sample	# 2	0.217	23:48
Breath Volume		3.453	23:48
Air Blank		0.000	23:48
Control Test		0.081	23:49
Air Blank		0.000	23:49
Diagnostics Check		OK	23:49

Data Download Date/Time: 05/16/2019 00:55

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHESSHER,SANDRA,W
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CHESSHER,S
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: MCARDLE

DATE/TIME OF TEST: 04/11/2019 01:34:56

Diagnostics Check		OK	01:27
Air Blank		0.000	01:28
Control Test		0.080	01:28
Air Blank		0.000	01:29
Subject Sample	# 1	0.239	01:30
Breath Volume		3.531	01:30
Air Blank		0.000	01:30
Air Blank		0.000	01:32
Subject Sample	# 2	0.233	01:33
Breath Volume		2.707	01:33
Air Blank		0.000	01:34
Control Test		0.081	01:34
Air Blank		0.000	01:34
Diagnostics Check		OK	01:34

Data Download Date/Time: 05/16/2019 00:55

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALHOUN,JOHN,A
BREATH TEST OPERATOR AGENCY: Astatula PD
ARRESTING OFFICER: SILVERIO,ALVIN
ARRESTING OFFICER AGENCY: Mascotte PD
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 17:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:57

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 04/11/2019 18:21:47

Diagnostics Check		OK	18:18	
Air Blank		0.000	18:19	
Control Test		0.080	18:19	
Air Blank		0.000	18:19	
Subject Sample	# 1		18:20	Subject Test Refused
Breath Volume		0.000	18:20	
Air Blank		0.000	18:20	
Control Test		0.081	18:21	
Air Blank		0.000	18:21	
Diagnostics Check		OK	18:21	

Data Download Date/Time: 05/16/2019 00:55

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALIJAGIC,SAMIR,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: ALIJAGIC,SAMIR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 06:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:41

SUBJECT'S LAST NAME: COURTNEY

DATE/TIME OF TEST: 04/12/2019 07:47:24

Diagnostics Check		OK	07:41
Air Blank		0.000	07:41
Control Test		0.080	07:41
Air Blank		0.000	07:42
Subject Sample	# 1	0.000	07:43
Breath Volume		2.664	07:43
Air Blank		0.000	07:43
Air Blank		0.000	07:45
Subject Sample	# 2	0.000	07:46
Breath Volume		3.207	07:46
Air Blank		0.000	07:46
Control Test		0.079	07:46
Air Blank		0.000	07:47
Diagnostics Check		OK	07:47

Data Download Date/Time: 05/16/2019 00:55

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPRAUN,ERICK,L
BREATH TEST OPERATOR AGENCY: Astatula PD
ARRESTING OFFICER: WRAY,SUSANNAH
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:48

SUBJECT'S LAST NAME: VERA

DATE/TIME OF TEST: 04/12/2019 21:15:31

Diagnostics Check		OK	21:12	
Air Blank		0.000	21:13	
Control Test		0.081	21:13	
Air Blank		0.000	21:13	
Subject Sample	# 1		21:14	Subject Test Refused
Breath Volume		0.000	21:14	
Air Blank		0.000	21:14	
Control Test		0.080	21:14	
Air Blank		0.000	21:15	
Diagnostics Check		OK	21:15	

Data Download Date/Time: 05/16/2019 00:55

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALHOUN,JOHN,A
BREATH TEST OPERATOR AGENCY: Astatula PD
ARRESTING OFFICER: CLARK,KURTIS
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: VENTURA

DATE/TIME OF TEST: 04/17/2019 00:13:18

Diagnostics Check		OK	00:06
Air Blank		0.000	00:07
Control Test		0.081	00:07
Air Blank		0.000	00:07
Subject Sample	# 1	0.211	00:08
Breath Volume		2.183	00:08
Air Blank		0.000	00:09
Air Blank		0.000	00:11
Subject Sample	# 2	0.214	00:11
Breath Volume		1.359	00:11
Air Blank		0.000	00:12
Control Test		0.081	00:12
Air Blank		0.000	00:13
Diagnostics Check		OK	00:13

Data Download Date/Time: 05/16/2019 00:55

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITCHER,DARRYL,P
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: PITCHER,DARRYL
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: PATTERSON

DATE/TIME OF TEST: 04/18/2019 03:37:18

Diagnostics Check		OK	03:30
Air Blank		0.000	03:31
Control Test		0.080	03:31
Air Blank		0.000	03:32
Subject Sample	# 1	0.208	03:32
Breath Volume		3.601	03:32
Air Blank		0.000	03:33
Air Blank		0.000	03:35
Subject Sample	# 2	0.201	03:35
Breath Volume		3.507	03:35
Air Blank		0.000	03:36
Control Test		0.081	03:36
Air Blank		0.000	03:37
Diagnostics Check		OK	03:37

Data Download Date/Time: 05/16/2019 00:55

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLLINS,MARVIN,D
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: COLLINS,MARVIN
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 05:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:35

SUBJECT'S LAST NAME: LOPEZ PEREZ

DATE/TIME OF TEST: 04/19/2019 07:07:08

Diagnostics Check		OK	06:55	
Air Blank		0.000	06:55	
Control Test		0.081	06:56	
Air Blank		0.000	06:56	
Subject Sample	# 1	0.155	06:59	Volume Not Met
Breath Volume		0.558	06:59	
Air Blank		0.000	07:00	
Air Blank		0.000	07:02	
Subject Sample	# 2	0.148	07:05	Volume Not Met
Breath Volume		0.046	07:05	
Air Blank		0.000	07:06	
Control Test		0.081	07:06	
Air Blank		0.000	07:07	
Diagnostics Check		OK	07:07	

Data Download Date/Time: 05/16/2019 00:55

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALHOUN,JOHN,A
BREATH TEST OPERATOR AGENCY: Astatula PD
ARRESTING OFFICER: CLARK,KURTIS
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: PANTAZIS

DATE/TIME OF TEST: 04/21/2019 01:27:34

Diagnostics Check		OK	01:21
Air Blank		0.000	01:21
Control Test		0.080	01:21
Air Blank		0.000	01:22
Subject Sample	# 1	0.207	01:23
Breath Volume		3.148	01:23
Air Blank		0.000	01:23
Air Blank		0.000	01:25
Subject Sample	# 2	0.205	01:26
Breath Volume		3.214	01:26
Air Blank		0.000	01:26
Control Test		0.081	01:26
Air Blank		0.000	01:27
Diagnostics Check		OK	01:27

Data Download Date/Time: 05/16/2019 00:55

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPRAUN,ERICK,L
BREATH TEST OPERATOR AGENCY: Astatula PD
ARRESTING OFFICER: CROW,STEVEN
ARRESTING OFFICER AGENCY: Umatilla PD
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 19:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:37

SUBJECT'S LAST NAME: HUTCHINSON

DATE/TIME OF TEST: 04/22/2019 20:01:04

Diagnostics Check		OK	19:58	
Air Blank		0.000	19:58	
Control Test		0.081	19:58	
Air Blank		0.000	19:59	
Subject Sample	# 1		19:59	Subject Test Refused
Breath Volume		0.000	19:59	
Air Blank		0.000	20:00	
Control Test		0.081	20:00	
Air Blank		0.000	20:00	
Diagnostics Check		OK	20:01	

Data Download Date/Time: 05/16/2019 00:55

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALSH,SHANNON,L
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: CHRISTIE,KEVIN
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 19:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: HARP

DATE/TIME OF TEST: 04/23/2019 22:13:32

Diagnostics Check		OK	22:09	
Air Blank		0.000	22:09	
Control Test		0.081	22:09	
Air Blank		0.000	22:10	
Subject Sample	# 1		22:12	Subject Test Refused
Breath Volume		0.000	22:12	
Air Blank		0.000	22:12	
Control Test		0.080	22:12	
Air Blank		0.000	22:13	
Diagnostics Check		OK	22:13	

Data Download Date/Time: 05/16/2019 00:55

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORMAN,GERALD,E
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CORMAN,GERALD
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: CHANTREE

DATE/TIME OF TEST: 04/25/2019 03:17:24

Diagnostics Check		OK	03:10
Air Blank		0.000	03:10
Control Test		0.080	03:11
Air Blank		0.000	03:11
Subject Sample	# 1	0.164	03:12
Breath Volume		1.707	03:12
Air Blank		0.000	03:12
Air Blank		0.000	03:14
Subject Sample	# 2	0.156	03:15
Breath Volume		1.832	03:15
Air Blank		0.000	03:16
Control Test		0.081	03:16
Air Blank		0.000	03:17
Diagnostics Check		OK	03:17

Data Download Date/Time: 05/16/2019 00:55

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALIJAGIC,SAMIR,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: MEREDITH,DENNISE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 11:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:16

SUBJECT'S LAST NAME: FOX

DATE/TIME OF TEST: 04/25/2019 12:42:18

Diagnostics Check		OK	12:36	
Air Blank		0.000	12:36	
Control Test		0.080	12:37	
Air Blank		0.000	12:37	
Subject Sample	# 1	0.225	12:40	Volume Not Met
Breath Volume		0.738	12:40	
Air Blank			12:41	Purge Fail
Air Blank		0.000	12:42	

Data Download Date/Time: 05/16/2019 00:56

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALIJAGIC,SAMIR,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: MEREDITH,DENNISE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 11:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:16

SUBJECT'S LAST NAME: FOX

DATE/TIME OF TEST: 04/25/2019 12:52:42

Diagnostics Check		OK	12:44
Air Blank		0.000	12:44
Control Test		0.078	12:45
Air Blank		0.000	12:45
Subject Sample	# 1	0.275	12:46
Breath Volume		1.441	12:46
Air Blank		0.000	12:47
Air Blank		0.000	12:49
Subject Sample	# 2	0.290	12:50
Breath Volume		2.117	12:50
Air Blank		0.000	12:51
Control Test		0.078	12:52
Air Blank		0.000	12:52
Diagnostics Check		OK	12:52

Data Download Date/Time: 05/16/2019 00:56

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALIJAGIC,SAMIR,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: ALIJAGIC,SAMIR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 12:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:56

SUBJECT'S LAST NAME: SKOLTE

DATE/TIME OF TEST: 04/26/2019 13:47:03

Diagnostics Check		OK	13:33	
Air Blank		0.000	13:33	
Control Test		0.080	13:34	
Air Blank		0.000	13:34	
Subject Sample	# 1	0.175	13:35	
Breath Volume		2.269	13:35	
Air Blank		0.000	13:36	
Air Blank		0.000	13:38	
Subject Sample	# 2	0.138	13:41	Volume Not Met
Breath Volume		1.000	13:41	
Air Blank		0.000	13:42	
Air Blank		0.000	13:43	
Subject Sample	# 3	0.165	13:45	
Breath Volume		2.214	13:45	
Air Blank		0.000	13:46	
Control Test		0.079	13:46	
Air Blank		0.000	13:46	
Diagnostics Check		OK	13:47	

Data Download Date/Time: 05/16/2019 00:56

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLOOM,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: BLOOM,NICHOLAS
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: PALENCIABARRERA

DATE/TIME OF TEST: 04/26/2019 22:47:34

Diagnostics Check		OK	22:31	
Air Blank		0.000	22:32	
Control Test		0.080	22:32	
Air Blank		0.000	22:33	
Subject Sample	# 1	0.146	22:36	Volume Not Met
Breath Volume		0.070	22:36	
Air Blank		0.000	22:37	
Air Blank		0.000	22:38	
Subject Sample	# 2	0.156	22:41	
Breath Volume		1.445	22:41	
Air Blank		0.000	22:42	
Air Blank		0.000	22:44	
Subject Sample	# 3	0.150	22:45	
Breath Volume		1.121	22:45	
Air Blank		0.000	22:46	
Control Test		0.079	22:46	
Air Blank		0.000	22:47	
Diagnostics Check		OK	22:47	

Data Download Date/Time: 05/16/2019 00:56

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,ETHAN,C
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: RAFFERTY,CHRIS
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: WIDEMAN

DATE/TIME OF TEST: 04/27/2019 02:36:31

Diagnostics Check		OK	02:30
Air Blank		0.000	02:30
Control Test		0.080	02:30
Air Blank		0.000	02:31
Subject Sample	# 1	0.055	02:32
Breath Volume		1.710	02:32
Air Blank		0.000	02:32
Air Blank		0.000	02:34
Subject Sample	# 2	0.054	02:34
Breath Volume		2.160	02:34
Air Blank		0.000	02:35
Control Test		0.080	02:35
Air Blank		0.000	02:36
Diagnostics Check		OK	02:36

Data Download Date/Time: 05/16/2019 00:56

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,ETHAN,C
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: KRUSE,KRIS
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: PIETROWICZ

DATE/TIME OF TEST: 04/27/2019 04:01:57

Diagnostics Check		OK	03:55
Air Blank		0.000	03:55
Control Test		0.080	03:56
Air Blank		0.000	03:56
Subject Sample	# 1	0.117	03:57
Breath Volume		1.832	03:57
Air Blank		0.000	03:58
Air Blank		0.000	03:59
Subject Sample	# 2	0.120	04:00
Breath Volume		1.718	04:00
Air Blank		0.000	04:01
Control Test		0.080	04:01
Air Blank		0.000	04:01
Diagnostics Check		OK	04:01

Data Download Date/Time: 05/16/2019 00:56

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,ETHAN,C
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: KRUSE,KRIS
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: SHIRLEY

DATE/TIME OF TEST: 04/27/2019 04:11:28

Diagnostics Check		OK	04:04
Air Blank		0.000	04:05
Control Test		0.079	04:05
Air Blank		0.000	04:05
Subject Sample	# 1	0.167	04:06
Breath Volume		2.132	04:06
Air Blank		0.000	04:07
Air Blank		0.000	04:09
Subject Sample	# 2	0.173	04:09
Breath Volume		2.437	04:09
Air Blank		0.000	04:10
Control Test		0.080	04:10
Air Blank		0.000	04:11
Diagnostics Check		OK	04:11

Data Download Date/Time: 05/16/2019 00:56

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,ETHAN,C
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: WALSH,SHANNON
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 04/27/2019 04:22:18

Diagnostics Check		OK	04:15
Air Blank		0.000	04:16
Control Test		0.079	04:16
Air Blank		0.000	04:17
Subject Sample	# 1	0.184	04:17
Breath Volume		3.808	04:17
Air Blank		0.000	04:18
Air Blank		0.000	04:20
Subject Sample	# 2	0.181	04:20
Breath Volume		3.843	04:20
Air Blank		0.000	04:21
Control Test		0.080	04:21
Air Blank		0.000	04:22
Diagnostics Check		OK	04:22

Data Download Date/Time: 05/16/2019 00:56

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLOOM,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: BLOOM,NICHOLAS
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: AUSTIN

DATE/TIME OF TEST: 04/27/2019 05:23:14

Diagnostics Check		OK	05:16
Air Blank		0.000	05:17
Control Test		0.080	05:17
Air Blank		0.000	05:18
Subject Sample	# 1	0.154	05:18
Breath Volume		2.769	05:18
Air Blank		0.000	05:19
Air Blank		0.000	05:21
Subject Sample	# 2	0.156	05:21
Breath Volume		2.546	05:21
Air Blank		0.000	05:22
Control Test		0.080	05:22
Air Blank		0.000	05:23
Diagnostics Check		OK	05:23

Data Download Date/Time: 05/16/2019 00:56

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRUSE,KRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: KRUSE,KRISTOPHER
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: ESTREMERA

DATE/TIME OF TEST: 04/28/2019 02:02:12

Diagnostics Check		OK	01:53	
Air Blank		0.000	01:54	
Control Test		0.080	01:54	
Air Blank		0.000	01:55	
Subject Sample	# 1		01:58	No Sample Provided
Breath Volume		0.000	01:58	
Air Blank		0.000	01:58	
Air Blank		0.000	02:00	
Subject Sample	# 2		02:00	Subject Test Refused
Breath Volume		0.000	02:00	
Air Blank		0.000	02:01	
Control Test		0.080	02:01	
Air Blank		0.000	02:02	
Diagnostics Check		OK	02:02	

Data Download Date/Time: 05/16/2019 00:56

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORMAN,GERALD,E
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CORMAN,GERALD
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: PIERCE

DATE/TIME OF TEST: 04/28/2019 03:07:56

Diagnostics Check		OK	03:01
Air Blank		0.000	03:02
Control Test		0.080	03:02
Air Blank		0.000	03:02
Subject Sample	# 1	0.156	03:03
Breath Volume		1.828	03:03
Air Blank		0.000	03:03
Air Blank		0.000	03:05
Subject Sample	# 2	0.166	03:06
Breath Volume		2.011	03:06
Air Blank		0.000	03:07
Control Test		0.081	03:07
Air Blank		0.000	03:07
Diagnostics Check		OK	03:07

Data Download Date/Time: 05/16/2019 00:56

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 04/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 16:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:47

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 04/06/2019 17:26:56

Diagnostics Check		OK	17:14	
Air Blank		0.000	17:14	
Control Test		0.080	17:15	
Air Blank		0.000	17:15	
Subject Sample	# 1	0.144	17:18	Volume Not Met
Breath Volume		0.996	17:18	
Air Blank		0.000	17:19	
Air Blank		0.000	17:21	
Subject Sample	# 2	0.137	17:22	
Breath Volume		1.394	17:22	
Air Blank		0.000	17:22	
Air Blank		0.000	17:24	
Subject Sample	# 3	0.130	17:25	
Breath Volume		1.113	17:25	
Air Blank		0.000	17:26	
Control Test		0.077	17:26	
Air Blank		0.000	17:26	
Diagnostics Check		OK	17:26	

Data Download Date/Time: 05/07/2019 09:49

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHERI,DIANE,M
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: ALLEN,FORREST
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 04/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 04/10/2019 03:53:16

Diagnostics Check		OK	03:50	
Air Blank		0.000	03:50	
Control Test		0.080	03:51	
Air Blank		0.000	03:51	
Subject Sample	# 1		03:51	Subject Test Refused
Breath Volume		0.000	03:51	
Air Blank		0.000	03:52	
Control Test		0.080	03:52	
Air Blank		0.000	03:53	
Diagnostics Check		OK	03:53	

Data Download Date/Time: 05/07/2019 09:49

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARAFA,YUSUF,A
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: ARAFA,YUSUF
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 04/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: PERMINOV

DATE/TIME OF TEST: 04/11/2019 04:22:36

Diagnostics Check		OK	04:11	
Air Blank		0.000	04:11	
Control Test		0.080	04:11	
Air Blank		0.000	04:12	
Subject Sample	# 1	0.230	04:15	Volume Not Met
Breath Volume		0.242	04:15	
Air Blank		0.000	04:16	
Air Blank		0.000	04:17	
Subject Sample	# 2	0.220	04:21	Volume Not Met
Breath Volume		0.566	04:21	
Air Blank		0.000	04:21	
Control Test		0.079	04:21	
Air Blank		0.000	04:22	
Diagnostics Check		OK	04:22	

Data Download Date/Time: 05/07/2019 09:49

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,JAMES,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: JONES,JAMES
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 04/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: BONDS

DATE/TIME OF TEST: 04/17/2019 22:49:58

Diagnostics Check		OK	22:43
Air Blank		0.000	22:43
Control Test		0.079	22:44
Air Blank		0.000	22:44
Subject Sample	# 1	0.201	22:45
Breath Volume		2.402	22:45
Air Blank		0.000	22:46
Air Blank		0.000	22:47
Subject Sample	# 2	0.187	22:48
Breath Volume		2.289	22:48
Air Blank		0.000	22:49
Control Test		0.079	22:49
Air Blank		0.000	22:49
Diagnostics Check		OK	22:49

Data Download Date/Time: 05/07/2019 09:49

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 04/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 22:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: WUEBBEN

DATE/TIME OF TEST: 04/19/2019 23:08:59

Diagnostics Check		OK	22:55	
Air Blank		0.000	22:56	
Control Test		0.080	22:56	
Air Blank		0.000	22:57	
Subject Sample	# 1		23:00	No Sample Provided
Breath Volume		0.000	23:00	
Air Blank		0.000	23:00	
Air Blank		0.000	23:02	
Subject Sample	# 2	0.171	23:04	
Breath Volume		1.210	23:04	
Air Blank		0.000	23:04	
Air Blank		0.000	23:06	
Subject Sample	# 3	0.187	23:07	
Breath Volume		1.621	23:07	
Air Blank		0.000	23:08	
Control Test		0.080	23:08	
Air Blank		0.000	23:08	
Diagnostics Check		OK	23:08	

Data Download Date/Time: 05/07/2019 09:49

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: WEBER,JEFFREY
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 04/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 04:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:58

SUBJECT'S LAST NAME: RIOS

DATE/TIME OF TEST: 04/20/2019 05:25:30

Diagnostics Check		OK	05:18
Air Blank		0.000	05:19
Control Test		0.081	05:19
Air Blank		0.000	05:19
Subject Sample	# 1	0.219	05:20
Breath Volume		4.582	05:20
Air Blank		0.000	05:21
Air Blank		0.000	05:23
Subject Sample	# 2	0.208	05:24
Breath Volume		2.972	05:24
Air Blank		0.000	05:24
Control Test		0.080	05:24
Air Blank		0.000	05:25
Diagnostics Check		OK	05:25

Data Download Date/Time: 05/07/2019 09:49

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,JAMES,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: JONES,JAMES
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 04/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 20:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:22

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 04/20/2019 21:05:45

Diagnostics Check		OK	20:59
Air Blank		0.000	21:00
Control Test		0.081	21:00
Air Blank		0.000	21:00
Subject Sample	# 1	0.155	21:01
Breath Volume		2.000	21:01
Air Blank		0.000	21:01
Air Blank		0.000	21:03
Subject Sample	# 2	0.152	21:04
Breath Volume		2.839	21:04
Air Blank		0.000	21:04
Control Test		0.080	21:05
Air Blank		0.000	21:05
Diagnostics Check		OK	21:05

Data Download Date/Time: 05/07/2019 09:49

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WURST,JOSHUA,D
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: WURST,JOSHUA
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 04/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: CRAWFORD

DATE/TIME OF TEST: 04/21/2019 02:37:32

Diagnostics Check		OK	02:31
Air Blank		0.000	02:31
Control Test		0.081	02:32
Air Blank		0.000	02:32
Subject Sample	# 1	0.235	02:33
Breath Volume		4.128	02:33
Air Blank		0.000	02:33
Air Blank		0.000	02:35
Subject Sample	# 2	0.225	02:36
Breath Volume		2.867	02:36
Air Blank		0.000	02:36
Control Test		0.079	02:36
Air Blank		0.000	02:37
Diagnostics Check		OK	02:37

Data Download Date/Time: 05/07/2019 09:49

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WURST,JOSHUA,D
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: WATSON,KEVIN
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 04/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 18:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:33

SUBJECT'S LAST NAME: MCNAIR

DATE/TIME OF TEST: 04/25/2019 19:45:27

Diagnostics Check		OK	19:42	
Air Blank		0.000	19:43	
Control Test		0.080	19:43	
Air Blank		0.000	19:43	
Subject Sample	# 1		19:44	Subject Test Refused
Breath Volume		0.000	19:44	
Air Blank		0.000	19:44	
Control Test		0.080	19:44	
Air Blank		0.000	19:45	
Diagnostics Check		OK	19:45	

Data Download Date/Time: 05/07/2019 09:49

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUEHLER,DEP,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BUEHLER,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 03/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:27

SUBJECT'S LAST NAME: LYNCH

DATE/TIME OF TEST: 04/13/2019 20:51:12

Diagnostics Check		OK	20:47	
Air Blank		0.000	20:48	
Control Test		0.078	20:48	
Air Blank		0.000	20:49	
Subject Sample	# 1		20:49	Slope Not Met
Breath Volume		1.976	20:49	
Air Blank		0.000	20:50	
Control Test		0.078	20:50	
Air Blank		0.000	20:51	
Diagnostics Check		OK	20:51	

Data Download Date/Time: 04/16/2019 08:13

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUEHLER,DEP,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BUEHLER,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 03/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:27

SUBJECT'S LAST NAME: LYNCH

DATE/TIME OF TEST: 04/13/2019 21:02:17

Diagnostics Check		OK	20:54
Air Blank		0.000	20:54
Control Test		0.078	20:54
Air Blank		0.000	20:55
Subject Sample	# 1	0.388	20:57
Breath Volume		4.058	20:57
Air Blank		0.000	20:58
Air Blank		0.000	20:59
Subject Sample	# 2	0.397	21:00
Breath Volume		3.429	21:00
Air Blank		0.000	21:01
Control Test		0.077	21:01
Air Blank		0.000	21:02
Diagnostics Check		OK	21:02

Data Download Date/Time: 04/16/2019 08:13

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

INSTRUMENT REGISTERED TO: BLOUNTSTOWN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,BRETT,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WAGNER,BRETT
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: CLARISSA

DATE/TIME OF TEST: 04/06/2019 02:37:21

Diagnostics Check		OK	02:30
Air Blank		0.000	02:31
Control Test		0.081	02:31
Air Blank		0.000	02:31
Subject Sample	# 1	0.234	02:32
Breath Volume		1.546	02:32
Air Blank		0.000	02:33
Air Blank		0.000	02:35
Subject Sample	# 2	0.222	02:35
Breath Volume		1.546	02:35
Air Blank		0.000	02:36
Control Test		0.080	02:36
Air Blank		0.000	02:37
Diagnostics Check		OK	02:37

Data Download Date/Time: 04/10/2019 12:15

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

INSTRUMENT REGISTERED TO: BLOUNTSTOWN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GENTRY,JONATHAN,P
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: GENTRY,JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 20:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 04/06/2019 21:33:29

Diagnostics Check		OK	21:25
Air Blank		0.000	21:25
Control Test		0.081	21:26
Air Blank		0.000	21:26
Subject Sample	# 1	0.125	21:28
Breath Volume		1.472	21:28
Air Blank		0.000	21:29
Air Blank		0.000	21:31
Subject Sample	# 2	0.122	21:31
Breath Volume		1.417	21:31
Air Blank		0.000	21:32
Control Test		0.081	21:32
Air Blank		0.000	21:33
Diagnostics Check		OK	21:33

Data Download Date/Time: 04/10/2019 12:15

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

INSTRUMENT REGISTERED TO: BLOUNTSTOWN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GENTRY,JONATHAN,P
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: GENTRY,JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 19:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:06

SUBJECT'S LAST NAME: CASS

DATE/TIME OF TEST: 04/10/2019 20:34:26

Diagnostics Check		OK	20:27
Air Blank		0.000	20:27
Control Test		0.082	20:28
Air Blank		0.000	20:28
Subject Sample	# 1	0.247	20:29
Breath Volume		2.316	20:29
Air Blank		0.000	20:30
Air Blank		0.000	20:32
Subject Sample	# 2	0.246	20:32
Breath Volume		1.582	20:32
Air Blank		0.000	20:33
Control Test		0.082	20:33
Air Blank		0.000	20:34
Diagnostics Check		OK	20:34

Data Download Date/Time: 05/05/2019 15:58

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

INSTRUMENT REGISTERED TO: BLOUNTSTOWN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINCOLN,TREY,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LINCOLN,TREY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 15:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:19

SUBJECT'S LAST NAME: CRANE

DATE/TIME OF TEST: 04/25/2019 15:55:39

Diagnostics Check		OK	15:48
Air Blank		0.000	15:48
Control Test		0.082	15:48
Air Blank		0.000	15:49
Subject Sample	# 1	0.195	15:50
Breath Volume		2.046	15:50
Air Blank		0.000	15:51
Air Blank		0.000	15:53
Subject Sample	# 2	0.186	15:54
Breath Volume		1.148	15:54
Air Blank		0.000	15:54
Control Test		0.082	15:55
Air Blank		0.000	15:55
Diagnostics Check		OK	15:55

Data Download Date/Time: 05/05/2019 15:58

Printed: 18 July 2019

COBRA ©1997-2014 CMI, Inc.