

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMENS, ADAM, J
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: CANNADAY, ZACHARY
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 03/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:00

SUBJECT'S LAST NAME: REYES NUNEZ

DATE/TIME OF TEST: 04/04/2019 08:27:00

Diagnostics Check	OK	08:20
Air Blank	0.000	08:21
Control Test	0.081	08:21
Air Blank	0.000	08:22
Subject Sample # 1	0.000	08:22
Breath Volume	5.027	08:22
Air Blank	0.000	08:22
Air Blank	0.000	08:24
Subject Sample # 2	0.000	08:25
Breath Volume	5.046	08:25
Air Blank	0.000	08:25
Control Test	0.080	08:26
Air Blank	0.000	08:26
Diagnostics Check	OK	08:27

Data Download Date/Time: 04/27/2019 20:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: KEE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 03/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: MELVIN

DATE/TIME OF TEST: 04/24/2019 01:02:37

Diagnostics Check	OK	00:46	
Air Blank	0.000	00:46	
Control Test	0.081	00:47	
Air Blank	0.000	00:47	
Subject Sample	# 1 0.239	00:50	Volume Not Met
Breath Volume	0.054	00:50	
Air Blank	0.000	00:51	
Air Blank	0.000	00:53	
Subject Sample	# 2 0.243	00:56	
Breath Volume	2.246	00:56	
Air Blank	0.000	00:56	
Air Blank	0.000	00:58	
Subject Sample	# 3 0.242	01:01	
Breath Volume	1.921	01:01	
Air Blank	0.000	01:01	
Control Test	0.083	01:02	
Air Blank	0.000	01:02	
Diagnostics Check	OK	01:02	

Data Download Date/Time: 04/27/2019 20:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: HARFORD,BRANDEN
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 03/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 04/25/2019 01:14:55

Diagnostics Check	OK	01:08
Air Blank	0.000	01:08
Control Test	0.081	01:09
Air Blank	0.000	01:09
Subject Sample # 1	0.000	01:10
Breath Volume	2.250	01:10
Air Blank	0.000	01:10
Air Blank	0.000	01:12
Subject Sample # 2	0.000	01:13
Breath Volume	1.835	01:13
Air Blank	0.000	01:14
Control Test	0.082	01:14
Air Blank	0.000	01:14
Diagnostics Check	OK	01:14

Data Download Date/Time: 04/27/2019 20:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: CRUZ,ALEXZANDER
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 04/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: HAMILTOON-TAGGART

DATE/TIME OF TEST: 04/26/2019 01:16:07

Diagnostics Check	OK	01:09
Air Blank	0.000	01:09
Control Test	0.081	01:10
Air Blank	0.000	01:10
Subject Sample # 1	0.149	01:11
Breath Volume	2.617	01:11
Air Blank	0.000	01:11
Air Blank	0.000	01:13
Subject Sample # 2	0.145	01:14
Breath Volume	2.710	01:14
Air Blank	0.000	01:15
Control Test	0.082	01:15
Air Blank	0.000	01:16
Diagnostics Check	OK	01:16

Data Download Date/Time: 04/27/2019 20:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: BURKE,ROBERT
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 04/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: SANABRIA

DATE/TIME OF TEST: 04/26/2019 03:26:14

Diagnostics Check	OK	03:16
Air Blank	0.000	03:17
Control Test	0.081	03:17
Air Blank	0.000	03:18
Subject Sample # 1	0.212	03:20
Breath Volume	2.113	03:20
Air Blank	0.000	03:21
Air Blank	0.000	03:23
Subject Sample # 2	0.216	03:24
Breath Volume	1.519	03:24
Air Blank	0.000	03:25
Control Test	0.081	03:25
Air Blank	0.000	03:26
Diagnostics Check	OK	03:26

Data Download Date/Time: 04/27/2019 20:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: KINKOPF,TIM
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 03/31/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: CHEVASCO

DATE/TIME OF TEST: 04/07/2019 04:48:23

Diagnostics Check	OK	04:42
Air Blank	0.000	04:42
Control Test	0.078	04:43
Air Blank	0.000	04:43
Subject Sample # 1	0.166	04:44
Breath Volume	1.308	04:44
Air Blank	0.000	04:44
Air Blank	0.000	04:46
Subject Sample # 2	0.163	04:46
Breath Volume	1.257	04:46
Air Blank	0.000	04:47
Control Test	0.078	04:47
Air Blank	0.000	04:48
Diagnostics Check	OK	04:48

Data Download Date/Time: 04/22/2019 21:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KRUPA,KRUPA,T

BREATH TEST OPERATOR AGENCY: Sanford PD

ARRESTING OFFICER: KRUPA,KYLE

ARRESTING OFFICER AGENCY: Sanford PD

LAST AGENCY INSPECTION: 03/31/2019

Date or Time Changed: Yes

DRY GAS STANDARD LOT #: 915310

DRY GAS EXPIRY DATE: 03/07/2021

VIOLATION CODE: DUI

ARREST TIME: 17:01

GENDER: Female

OBSERVATION PERIOD BEGAN: 18:08

SUBJECT'S LAST NAME: HARTMAN

DATE/TIME OF TEST: 04/11/2019 18:51:33

Diagnostics Check	OK	18:40
Air Blank	0.000	18:40
Control Test	0.079	18:41
Air Blank	0.000	18:41
Subject Sample # 1	0.250	18:43
Breath Volume	2.890	18:43
Air Blank	0.000	18:43
Air Blank	0.000	18:45
Subject Sample # 2	0.227	18:46
Breath Volume	1.625	18:46
Air Blank	0.000	18:46
Air Blank	0.000	18:48
Subject Sample # 3	0.223	18:50
Breath Volume	1.285	18:50
Air Blank	0.000	18:50
Control Test	0.078	18:50
Air Blank	0.000	18:51
Diagnostics Check	OK	18:51

Data Download Date/Time: 04/22/2019 21:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUDBRINK,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SUDBRINK,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: MOSCA

DATE/TIME OF TEST: 04/01/2019 00:15:42

Diagnostics Check	OK	00:09
Air Blank	0.000	00:10
Control Test	0.080	00:10
Air Blank	0.000	00:10
Subject Sample # 1	0.178	00:11
Breath Volume	2.488	00:11
Air Blank	0.000	00:11
Air Blank	0.000	00:13
Subject Sample # 2	0.183	00:14
Breath Volume	3.054	00:14
Air Blank	0.000	00:14
Control Test	0.081	00:15
Air Blank	0.000	00:15
Diagnostics Check	OK	00:15

Data Download Date/Time: 04/18/2019 03:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CAMPION,KRISTIE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: OHLWEILER

DATE/TIME OF TEST: 04/02/2019 21:57:08

Diagnostics Check	OK	21:50
Air Blank	0.000	21:50
Control Test	0.081	21:51
Air Blank	0.000	21:51
Subject Sample # 1	0.164	21:52
Breath Volume	1.746	21:52
Air Blank	0.000	21:52
Air Blank	0.000	21:54
Subject Sample # 2	0.174	21:55
Breath Volume	1.644	21:55
Air Blank	0.000	21:56
Control Test	0.081	21:56
Air Blank	0.000	21:57
Diagnostics Check	OK	21:57

Data Download Date/Time: 04/18/2019 03:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MUNOZ,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: BLAIR

DATE/TIME OF TEST: 04/02/2019 22:47:14

Diagnostics Check	OK	22:44	
Air Blank	0.000	22:44	
Control Test	0.082	22:45	
Air Blank	0.000	22:45	
Subject Sample	# 1	22:45	Subject Test Refused
Breath Volume	0.000	22:45	
Air Blank	0.000	22:46	
Control Test	0.081	22:46	
Air Blank	0.000	22:47	
Diagnostics Check	OK	22:47	

Data Download Date/Time: 04/18/2019 03:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,NELSON,X
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PUPKE,LUKE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: DIAZ-CAMPOS

DATE/TIME OF TEST: 04/03/2019 02:35:24

Diagnostics Check	OK	02:32	
Air Blank	0.000	02:32	
Control Test	0.081	02:32	
Air Blank	0.000	02:33	
Subject Sample	# 1	02:34	Subject Test Refused
Breath Volume	0.000	02:34	
Air Blank	0.000	02:34	
Control Test	0.081	02:34	
Air Blank	0.000	02:35	
Diagnostics Check	OK	02:35	

Data Download Date/Time: 04/18/2019 03:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, WILLIAM, T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: RICHO

DATE/TIME OF TEST: 04/04/2019 03:26:52

Diagnostics Check	OK	03:20
Air Blank	0.000	03:20
Control Test	0.081	03:21
Air Blank	0.000	03:21
Subject Sample # 1	0.109	03:22
Breath Volume	1.437	03:22
Air Blank	0.000	03:23
Air Blank	0.000	03:24
Subject Sample # 2	0.104	03:25
Breath Volume	2.621	03:25
Air Blank	0.000	03:25
Control Test	0.082	03:26
Air Blank	0.000	03:26
Diagnostics Check	OK	03:26

Data Download Date/Time: 04/18/2019 03:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEELER,BRYAN,C
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LAMBIRD,DAVID
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 07:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:35

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 04/04/2019 08:59:13

Diagnostics Check	OK	08:55	
Air Blank	0.000	08:56	
Control Test	0.082	08:56	
Air Blank	0.000	08:57	
Subject Sample	# 1	08:57	RFI Detect
Breath Volume	0.000	08:57	
Air Blank	0.000	08:58	
Control Test	0.082	08:58	
Air Blank	0.000	08:59	
Diagnostics Check	OK	08:59	

Data Download Date/Time: 04/18/2019 03:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,AARON,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BEHRENSHOUSER,RICHARD
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: HOOD

DATE/TIME OF TEST: 04/04/2019 18:46:53

Diagnostics Check	OK	18:40
Air Blank	0.000	18:40
Control Test	0.081	18:41
Air Blank	0.000	18:41
Subject Sample # 1	0.000	18:42
Breath Volume	1.492	18:42
Air Blank	0.000	18:42
Air Blank	0.000	18:44
Subject Sample # 2	0.000	18:45
Breath Volume	1.199	18:45
Air Blank	0.000	18:45
Control Test	0.082	18:46
Air Blank	0.000	18:46
Diagnostics Check	OK	18:46

Data Download Date/Time: 04/18/2019 03:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICOTTILLI, JOSHUA, A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI, JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 04/05/2019 03:23:25

Diagnostics Check	OK	03:15	
Air Blank	0.000	03:15	
Control Test	0.081	03:15	
Air Blank	0.000	03:16	
Subject Sample	# 1	03:19	No Sample Provided
Breath Volume	0.000	03:19	
Air Blank	0.000	03:20	
Air Blank	0.000	03:21	
Subject Sample	# 2	03:22	Subject Test Refused
Breath Volume	0.000	03:22	
Air Blank	0.000	03:22	
Control Test	0.081	03:22	
Air Blank	0.000	03:23	
Diagnostics Check	OK	03:23	

Data Download Date/Time: 04/18/2019 03:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE,DEPUTY WILLIAM,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,DEPUTY L.
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: AZEMAN

DATE/TIME OF TEST: 04/08/2019 01:20:12

Diagnostics Check	OK	01:12
Air Blank	0.000	01:12
Control Test	0.081	01:13
Air Blank	0.000	01:13
Subject Sample # 1	0.214	01:15
Breath Volume	1.628	01:15
Air Blank	0.000	01:16
Air Blank	0.000	01:17
Subject Sample # 2	0.201	01:18
Breath Volume	1.437	01:18
Air Blank	0.000	01:19
Control Test	0.081	01:19
Air Blank	0.000	01:20
Diagnostics Check	OK	01:20

Data Download Date/Time: 04/18/2019 03:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WORTHINGTON,JUSTIN,C
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WORTHINGTON,JUSTIN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: MEDINA

DATE/TIME OF TEST: 04/08/2019 05:13:46

Diagnostics Check	OK	05:06
Air Blank	0.000	05:06
Control Test	0.081	05:06
Air Blank	0.000	05:07
Subject Sample # 1	0.012	05:08
Breath Volume	1.273	05:08
Air Blank	0.000	05:09
Air Blank	0.000	05:11
Subject Sample # 2	0.011	05:12
Breath Volume	1.238	05:12
Air Blank	0.000	05:12
Control Test	0.081	05:13
Air Blank	0.000	05:13
Diagnostics Check	OK	05:13

Data Download Date/Time: 04/18/2019 03:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICOTTILLI, JOSHUA, A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI, JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: VELLINGA

DATE/TIME OF TEST: 04/11/2019 02:56:17

Diagnostics Check	OK	02:42	
Air Blank	0.000	02:43	
Control Test	0.082	02:43	
Air Blank	0.000	02:43	
Subject Sample	# 1	02:47	No Sample Provided
Breath Volume	0.000	02:47	
Air Blank	0.000	02:47	
Air Blank	0.000	02:49	
Subject Sample	# 2	0.187	02:51
Breath Volume	1.406	02:51	
Air Blank	0.000	02:52	
Air Blank	0.000	02:54	
Subject Sample	# 3	0.190	02:54
Breath Volume	1.410	02:54	
Air Blank	0.000	02:55	
Control Test	0.082	02:55	
Air Blank	0.000	02:56	
Diagnostics Check	OK	02:56	

Data Download Date/Time: 04/18/2019 03:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: MAHONY

DATE/TIME OF TEST: 04/12/2019 01:49:36

Diagnostics Check	OK	01:46	
Air Blank	0.000	01:47	
Control Test	0.082	01:47	
Air Blank	0.000	01:48	
Subject Sample	# 1	01:48	Subject Test Refused
Breath Volume	0.000	01:48	
Air Blank	0.000	01:48	
Control Test	0.082	01:49	
Air Blank	0.000	01:49	
Diagnostics Check	OK	01:49	

Data Download Date/Time: 04/18/2019 03:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICOTTILLI, JOSHUA, A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI, JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: TUMMINIA

DATE/TIME OF TEST: 04/12/2019 02:28:28

Diagnostics Check	OK	02:25	
Air Blank	0.000	02:26	
Control Test	0.081	02:26	
Air Blank	0.000	02:26	
Subject Sample	# 1	02:27	Subject Test Refused
Breath Volume	0.000	02:27	
Air Blank	0.000	02:27	
Control Test	0.082	02:27	
Air Blank	0.000	02:28	
Diagnostics Check	OK	02:28	

Data Download Date/Time: 04/18/2019 03:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: HENDON

DATE/TIME OF TEST: 04/13/2019 03:10:38

Diagnostics Check	OK	03:07	
Air Blank	0.000	03:08	
Control Test	0.081	03:08	
Air Blank	0.000	03:08	
Subject Sample	# 1	03:09	Subject Test Refused
Breath Volume	0.000	03:09	
Air Blank	0.000	03:09	
Control Test	0.081	03:10	
Air Blank	0.000	03:10	
Diagnostics Check	OK	03:10	

Data Download Date/Time: 04/18/2019 03:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZSIDO,D
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: GEEL

DATE/TIME OF TEST: 04/13/2019 04:23:23

Diagnostics Check	OK	04:15
Air Blank	0.000	04:16
Control Test	0.082	04:16
Air Blank	0.000	04:16
Subject Sample # 1	0.223	04:18
Breath Volume	2.136	04:18
Air Blank	0.000	04:19
Air Blank	0.000	04:21
Subject Sample # 2	0.223	04:21
Breath Volume	2.281	04:21
Air Blank	0.000	04:22
Control Test	0.082	04:22
Air Blank	0.000	04:23
Diagnostics Check	OK	04:23

Data Download Date/Time: 04/18/2019 03:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE,DEPUTY WILLIAM,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,DEPUTY L.
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: ROSSI

DATE/TIME OF TEST: 04/16/2019 02:00:19

Diagnostics Check	OK	01:53
Air Blank	0.000	01:53
Control Test	0.081	01:53
Air Blank	0.000	01:54
Subject Sample # 1	0.082	01:55
Breath Volume	2.484	01:55
Air Blank	0.000	01:56
Air Blank	0.000	01:58
Subject Sample # 2	0.081	01:58
Breath Volume	2.046	01:58
Air Blank	0.000	01:59
Control Test	0.082	01:59
Air Blank	0.000	02:00
Diagnostics Check	OK	02:00

Data Download Date/Time: 04/18/2019 03:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRIDGES,DARREN,K
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BRIDGES,DARREN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 13:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:32

SUBJECT'S LAST NAME: KEEBLE

DATE/TIME OF TEST: 04/16/2019 14:46:44

Diagnostics Check	OK	14:43	
Air Blank	0.000	14:44	
Control Test	0.081	14:44	
Air Blank	0.000	14:45	
Subject Sample	# 1	14:45	Subject Test Refused
Breath Volume	0.000	14:45	
Air Blank	0.000	14:45	
Control Test	0.081	14:46	
Air Blank	0.000	14:46	
Diagnostics Check	OK	14:46	

Data Download Date/Time: 04/18/2019 03:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REEVES, MARK, O
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN, BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 11:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:06

SUBJECT'S LAST NAME: DAUGHERTY

DATE/TIME OF TEST: 04/20/2019 13:33:02

Diagnostics Check	OK	13:26
Air Blank	0.000	13:27
Control Test	0.081	13:27
Air Blank	0.000	13:28
Subject Sample # 1	0.000	13:28
Breath Volume	2.621	13:28
Air Blank	0.000	13:28
Air Blank	0.000	13:30
Subject Sample # 2	0.000	13:31
Breath Volume	2.718	13:31
Air Blank	0.000	13:32
Control Test	0.082	13:32
Air Blank	0.000	13:32
Diagnostics Check	OK	13:33

Data Download Date/Time: 05/11/2019 22:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REEVES,MARK,O
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MORRISON,QUINN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 14:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:43

SUBJECT'S LAST NAME: PETRINA

DATE/TIME OF TEST: 04/20/2019 15:11:06

Diagnostics Check	OK	15:04
Air Blank	0.000	15:04
Control Test	0.081	15:05
Air Blank	0.000	15:05
Subject Sample # 1	0.184	15:06
Breath Volume	1.898	15:06
Air Blank	0.000	15:07
Air Blank	0.000	15:08
Subject Sample # 2	0.184	15:09
Breath Volume	1.878	15:09
Air Blank	0.000	15:10
Control Test	0.079	15:10
Air Blank	0.000	15:11
Diagnostics Check	OK	15:11

Data Download Date/Time: 05/11/2019 22:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICOTTILLI, JOSHUA, A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI, JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: EARLS

DATE/TIME OF TEST: 04/20/2019 22:45:18

Diagnostics Check	OK	22:41	
Air Blank	0.000	22:41	
Control Test	0.081	22:42	
Air Blank	0.000	22:42	
Subject Sample	# 1	22:43	Subject Test Refused
Breath Volume	0.000	22:43	
Air Blank	0.000	22:44	
Control Test	0.081	22:44	
Air Blank	0.000	22:45	
Diagnostics Check	OK	22:45	

Data Download Date/Time: 05/11/2019 22:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: REYNOLDS

DATE/TIME OF TEST: 04/21/2019 02:43:23

Diagnostics Check	OK	02:36
Air Blank	0.000	02:36
Control Test	0.081	02:36
Air Blank	0.000	02:37
Subject Sample # 1	0.000	02:37
Breath Volume	1.296	02:37
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample # 2	0.000	02:42
Breath Volume	1.460	02:42
Air Blank	0.000	02:42
Control Test	0.082	02:42
Air Blank	0.000	02:43
Diagnostics Check	OK	02:43

Data Download Date/Time: 05/11/2019 22:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICOTTILLI, JOSHUA, A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI, JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 04/21/2019 03:31:31

Diagnostics Check	OK	03:22	
Air Blank	0.000	03:23	
Control Test	0.081	03:23	
Air Blank	0.000	03:23	
Subject Sample	# 1	03:27	No Sample Provided
Breath Volume	0.000	03:27	
Air Blank	0.000	03:27	
Air Blank	0.000	03:29	
Subject Sample	# 2	03:30	Subject Test Refused
Breath Volume	0.000	03:30	
Air Blank	0.000	03:30	
Control Test	0.081	03:30	
Air Blank	0.000	03:31	
Diagnostics Check	OK	03:31	

Data Download Date/Time: 05/11/2019 22:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE,DEPUTY WILLIAM,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FICK III,DEPUTY G.
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: CHICCINE

DATE/TIME OF TEST: 04/21/2019 20:06:00

Diagnostics Check	OK	20:03	
Air Blank	0.000	20:03	
Control Test	0.082	20:03	
Air Blank	0.000	20:04	
Subject Sample	# 1	20:04	Subject Test Refused
Breath Volume	0.000	20:04	
Air Blank	0.000	20:05	
Control Test	0.081	20:05	
Air Blank	0.000	20:05	
Diagnostics Check	OK	20:06	

Data Download Date/Time: 05/11/2019 22:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICOTTILLI, JOSHUA, A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI, JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:23

SUBJECT'S LAST NAME: RODDY

DATE/TIME OF TEST: 04/21/2019 21:00:55

Diagnostics Check	OK	20:52
Air Blank	0.000	20:53
Control Test	0.082	20:53
Air Blank	0.000	20:53
Subject Sample # 1	0.000	20:54
Breath Volume	1.539	20:54
Air Blank	0.000	20:55
Air Blank	0.000	20:57
Subject Sample # 2	0.000	20:59
Breath Volume	1.527	20:59
Air Blank	0.000	21:00
Control Test	0.081	21:00
Air Blank	0.000	21:00
Diagnostics Check	OK	21:00

Data Download Date/Time: 05/11/2019 22:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE,DEPUTY WILLIAM,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WILTSE,DEPUTY WILLIAM
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: BUI

DATE/TIME OF TEST: 04/21/2019 23:44:41

Diagnostics Check	OK	23:38
Air Blank	0.000	23:38
Control Test	0.082	23:39
Air Blank	0.000	23:39
Subject Sample # 1	0.155	23:40
Breath Volume	2.699	23:40
Air Blank	0.000	23:40
Air Blank	0.000	23:42
Subject Sample # 2	0.161	23:43
Breath Volume	2.742	23:43
Air Blank	0.000	23:43
Control Test	0.082	23:44
Air Blank	0.000	23:44
Diagnostics Check	OK	23:44

Data Download Date/Time: 05/11/2019 22:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE,DEPUTY WILLIAM,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,DEPUTY L.
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: MC KELVEY

DATE/TIME OF TEST: 04/22/2019 00:22:28

Diagnostics Check	OK	00:19	
Air Blank	0.000	00:19	
Control Test	0.082	00:20	
Air Blank	0.000	00:20	
Subject Sample	# 1	00:21	Subject Test Refused
Breath Volume	0.000	00:21	
Air Blank	0.000	00:21	
Control Test	0.081	00:21	
Air Blank	0.000	00:22	
Diagnostics Check	OK	00:22	

Data Download Date/Time: 05/11/2019 22:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE,DEPUTY WILLIAM,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WILTSE,DEPUTY WILLIAM
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 04/22/2019 02:28:08

Diagnostics Check	OK	02:21
Air Blank	0.000	02:21
Control Test	0.082	02:21
Air Blank	0.000	02:22
Subject Sample # 1	0.085	02:23
Breath Volume	1.929	02:23
Air Blank	0.000	02:24
Air Blank	0.000	02:25
Subject Sample # 2	0.081	02:26
Breath Volume	2.367	02:26
Air Blank	0.000	02:27
Control Test	0.083	02:27
Air Blank	0.000	02:28
Diagnostics Check	OK	02:28

Data Download Date/Time: 05/11/2019 22:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICOTTILLI, JOSHUA, A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI, JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: NORRIS

DATE/TIME OF TEST: 04/26/2019 00:40:34

Diagnostics Check	OK	00:31
Air Blank	0.000	00:32
Control Test	0.081	00:32
Air Blank	0.000	00:32
Subject Sample # 1	0.181	00:35
Breath Volume	1.593	00:35
Air Blank	0.000	00:35
Air Blank	0.000	00:37
Subject Sample # 2	0.184	00:38
Breath Volume	1.722	00:38
Air Blank	0.000	00:39
Control Test	0.082	00:39
Air Blank	0.000	00:40
Diagnostics Check	OK	00:40

Data Download Date/Time: 05/11/2019 22:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONGMIRE,BENJAMIN,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: LONGMIRE,BENJAMIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: BEYER

DATE/TIME OF TEST: 04/26/2019 03:47:26

Diagnostics Check	OK	03:40
Air Blank	0.000	03:41
Control Test	0.081	03:41
Air Blank	0.000	03:42
Subject Sample # 1	0.238	03:42
Breath Volume	2.382	03:42
Air Blank	0.000	03:43
Air Blank	0.000	03:45
Subject Sample # 2	0.235	03:45
Breath Volume	3.773	03:45
Air Blank	0.000	03:46
Control Test	0.081	03:46
Air Blank	0.000	03:47
Diagnostics Check	OK	03:47

Data Download Date/Time: 05/11/2019 22:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTS,C,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BATTS,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 04/27/2019 01:12:57

Diagnostics Check	OK	01:05
Air Blank	0.000	01:06
Control Test	0.082	01:06
Air Blank	0.000	01:06
Subject Sample # 1	0.188	01:08
Breath Volume	1.886	01:08
Air Blank	0.000	01:09
Air Blank	0.000	01:10
Subject Sample # 2	0.186	01:11
Breath Volume	1.964	01:11
Air Blank	0.000	01:12
Control Test	0.081	01:12
Air Blank	0.000	01:12
Diagnostics Check	OK	01:12

Data Download Date/Time: 05/11/2019 22:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICOTTILLI,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI,JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: COLSON

DATE/TIME OF TEST: 04/27/2019 02:59:44

Diagnostics Check	OK	02:56	
Air Blank	0.000	02:57	
Control Test	0.081	02:57	
Air Blank	0.000	02:58	
Subject Sample	# 1	02:58	RFI Detect
Breath Volume	0.000	02:58	
Air Blank	0.000	02:58	
Control Test	0.081	02:59	
Air Blank	0.000	02:59	
Diagnostics Check	OK	02:59	

Data Download Date/Time: 05/11/2019 22:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICOTTILLI,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI,JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: COLSON

DATE/TIME OF TEST: 04/27/2019 03:17:33

Diagnostics Check	OK	03:02	
Air Blank	0.000	03:03	
Control Test	0.083	03:03	
Air Blank	0.000	03:04	
Subject Sample	# 1 0.109	03:07	Volume Not Met
Breath Volume	0.222	03:07	
Air Blank	0.000	03:07	
Air Blank	0.000	03:09	
Subject Sample	# 2 0.115	03:12	
Breath Volume	1.265	03:12	
Air Blank	0.000	03:12	
Air Blank	0.000	03:14	
Subject Sample	# 3 0.117	03:16	
Breath Volume	1.332	03:16	
Air Blank	0.000	03:16	
Control Test	0.081	03:16	
Air Blank	0.000	03:17	
Diagnostics Check	OK	03:17	

Data Download Date/Time: 05/11/2019 22:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 04/27/2019 03:52:36

Diagnostics Check	OK	03:45
Air Blank	0.000	03:46
Control Test	0.082	03:46
Air Blank	0.000	03:47
Subject Sample # 1	0.072	03:47
Breath Volume	2.730	03:47
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 2	0.070	03:51
Breath Volume	2.101	03:51
Air Blank	0.000	03:51
Control Test	0.081	03:52
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time: 05/11/2019 22:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONGMIRE,BENJAMIN,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: LONGMIRE,BENJAMIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 11:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: ZACKERU

DATE/TIME OF TEST: 04/28/2019 00:43:25

Diagnostics Check	OK	00:37
Air Blank	0.000	00:37
Control Test	0.081	00:37
Air Blank	0.000	00:38
Subject Sample # 1	0.176	00:38
Breath Volume	3.765	00:38
Air Blank	0.000	00:39
Air Blank	0.000	00:41
Subject Sample # 2	0.175	00:41
Breath Volume	4.007	00:41
Air Blank	0.000	00:42
Control Test	0.080	00:42
Air Blank	0.000	00:43
Diagnostics Check	OK	00:43

Data Download Date/Time: 05/11/2019 22:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTS,C,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BATTS,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: PEARSON

DATE/TIME OF TEST: 04/28/2019 01:49:19

Diagnostics Check	OK	01:42
Air Blank	0.000	01:43
Control Test	0.080	01:43
Air Blank	0.000	01:44
Subject Sample # 1	0.153	01:44
Breath Volume	3.542	01:44
Air Blank	0.000	01:45
Air Blank	0.000	01:47
Subject Sample # 2	0.151	01:47
Breath Volume	3.285	01:47
Air Blank	0.000	01:48
Control Test	0.081	01:48
Air Blank	0.000	01:49
Diagnostics Check	OK	01:49

Data Download Date/Time: 05/11/2019 22:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE,DEPUTY WILLIAM,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DIEBEL,DEPUTY J.
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:54

SUBJECT'S LAST NAME: PIERCE

DATE/TIME OF TEST: 04/30/2019 19:24:09

Diagnostics Check	OK	19:17
Air Blank	0.000	19:18
Control Test	0.081	19:18
Air Blank	0.000	19:18
Subject Sample # 1	0.244	19:19
Breath Volume	2.539	19:19
Air Blank	0.000	19:20
Air Blank	0.000	19:22
Subject Sample # 2	0.234	19:22
Breath Volume	1.726	19:22
Air Blank	0.000	19:23
Control Test	0.080	19:23
Air Blank	0.000	19:24
Diagnostics Check	OK	19:24

Data Download Date/Time: 05/11/2019 22:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE,DEPUTY WILLIAM,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DUCKERS,DEPUTY R.
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: EAMES

DATE/TIME OF TEST: 04/30/2019 23:42:24

Diagnostics Check	OK	23:25	
Air Blank	0.000	23:25	
Control Test	0.082	23:26	
Air Blank	0.000	23:26	
Subject Sample	# 1 0.192	23:29	Volume Not Met
Breath Volume	0.035	23:29	
Air Blank	0.000	23:30	
Air Blank	0.000	23:32	
Subject Sample	# 2 0.179	23:35	
Breath Volume	1.691	23:35	
Air Blank	0.000	23:35	
Air Blank	0.000	23:37	
Subject Sample	# 3 0.190	23:40	Volume Not Met
Breath Volume	0.707	23:40	
Air Blank	0.000	23:41	
Control Test	0.080	23:41	
Air Blank	0.000	23:42	
Diagnostics Check	OK	23:42	

Data Download Date/Time: 05/11/2019 22:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE,DEPUTY WILLIAM,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DUCKERS,DEPUTY R.
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: EAMES

DATE/TIME OF TEST: 04/30/2019 23:53:49

Diagnostics Check	OK	23:46
Air Blank	0.000	23:46
Control Test	0.080	23:46
Air Blank	0.000	23:47
Subject Sample # 1	0.178	23:48
Breath Volume	2.062	23:48
Air Blank	0.000	23:49
Air Blank	0.000	23:50
Subject Sample # 2	0.176	23:52
Breath Volume	1.679	23:52
Air Blank	0.000	23:52
Control Test	0.080	23:53
Air Blank	0.000	23:53
Diagnostics Check	OK	23:53

Data Download Date/Time: 05/11/2019 22:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORTON,GREG,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ARLINGTON,NICHOLAS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 10:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:40

SUBJECT'S LAST NAME: BISHOP

DATE/TIME OF TEST: 04/13/2019 12:07:25

Diagnostics Check	OK	12:00
Air Blank	0.000	12:01
Control Test	0.080	12:01
Air Blank	0.000	12:01
Subject Sample # 1	0.000	12:03
Breath Volume	1.703	12:03
Air Blank	0.000	12:03
Air Blank	0.000	12:05
Subject Sample # 2	0.000	12:06
Breath Volume	1.687	12:06
Air Blank	0.000	12:06
Control Test	0.079	12:06
Air Blank	0.000	12:07
Diagnostics Check	OK	12:07

Data Download Date/Time: 04/17/2019 21:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHRISTIE, MATTHEW, J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHRISTIE, MATTHEW
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/17/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: PAINTER

DATE/TIME OF TEST: 04/24/2019 01:02:51

Diagnostics Check	OK	01:01	
Air Blank	0.000	01:01	
Control Test	0.080	01:01	
Air Blank	0.000	01:02	
Subject Sample	# 1	01:02	RFI Detect
Breath Volume	0.000	01:02	
Air Blank		01:02	RFI Detect
Air Blank		01:02	RFI Detect

Data Download Date/Time: 05/11/2019 21:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHRISTIE, MATTHEW, J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHRISTIE, MATTHEW
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/17/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: PAINTER

DATE/TIME OF TEST: 04/24/2019 01:08:15

Diagnostics Check	OK	01:05	
Air Blank	0.000	01:05	
Control Test	0.079	01:06	
Air Blank	0.000	01:06	
Subject Sample	# 1	01:06	RFI Detect
Breath Volume	0.000	01:06	
Air Blank	0.000	01:07	
Control Test	0.080	01:07	
Air Blank	0.000	01:08	
Diagnostics Check	OK	01:08	

Data Download Date/Time: 05/11/2019 21:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHRISTIE, MATTHEW, J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHRISTIE, MATTHEW
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/17/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: PAINTER

DATE/TIME OF TEST: 04/24/2019 01:20:08

Diagnostics Check	OK	01:12
Air Blank	0.000	01:12
Control Test	0.080	01:13
Air Blank	0.000	01:13
Subject Sample # 1	0.153	01:14
Breath Volume	1.296	01:14
Air Blank	0.000	01:14
Air Blank	0.000	01:16
Subject Sample # 2	0.154	01:18
Breath Volume	2.097	01:18
Air Blank	0.000	01:19
Control Test	0.080	01:19
Air Blank	0.000	01:20
Diagnostics Check	OK	01:20

Data Download Date/Time: 05/11/2019 21:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

INSTRUMENT REGISTERED TO: INDIALANTIC P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOURNIER,TIMOTHY,R
BREATH TEST OPERATOR AGENCY: Indialantic PD
ARRESTING OFFICER: HINES,MICHAEL
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 03/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: BARRIENTOS

DATE/TIME OF TEST: 04/17/2019 03:33:12

Diagnostics Check	OK	03:21	
Air Blank	0.000	03:22	
Control Test	0.079	03:22	
Air Blank	0.000	03:23	
Subject Sample	# 1	03:26	No Sample Provided
Breath Volume	0.000	03:26	
Air Blank	0.000	03:26	
Air Blank	0.000	03:28	
Subject Sample	# 2	03:31	No Sample Provided
Breath Volume	0.000	03:31	
Air Blank	0.000	03:32	
Control Test	0.080	03:32	
Air Blank	0.000	03:33	
Diagnostics Check	OK	03:33	

Data Download Date/Time: 05/02/2019 07:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

INSTRUMENT REGISTERED TO: INDIALANTIC P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOURNIER,TIMOTHY,R
BREATH TEST OPERATOR AGENCY: Indialantic PD
ARRESTING OFFICER: FOURNIER,TIMOTHY
ARRESTING OFFICER AGENCY: Indialantic PD
LAST AGENCY INSPECTION: 03/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 04/19/2019 03:35:29

Diagnostics Check	OK	03:28
Air Blank	0.000	03:28
Control Test	0.080	03:28
Air Blank	0.000	03:29
Subject Sample # 1	0.250	03:30
Breath Volume	4.683	03:30
Air Blank	0.000	03:31
Air Blank	0.000	03:32
Subject Sample # 2	0.245	03:33
Breath Volume	4.535	03:33
Air Blank	0.000	03:34
Control Test	0.079	03:34
Air Blank	0.000	03:35
Diagnostics Check	OK	03:35

Data Download Date/Time: 05/02/2019 07:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAGER,KEVIN,M
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: HAGER,KEVIN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: PROB/PAROL
ARREST TIME: 08:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:00

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 04/09/2019 08:40:24

Diagnostics Check	OK	08:37	
Air Blank	0.000	08:37	
Control Test	0.081	08:38	
Air Blank	0.000	08:38	
Subject Sample	# 1	08:38	Interferent Detect
Breath Volume	1.156	08:38	
Air Blank	0.000	08:39	
Control Test	0.084	08:39	
Air Blank	0.000	08:40	
Diagnostics Check	OK	08:40	

Data Download Date/Time: 05/30/2019 15:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KALLMEYER,DANIEL,K
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: N/A,N/A
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: PROB/PAROL
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:00

SUBJECT'S LAST NAME: HITLEY

DATE/TIME OF TEST: 04/09/2019 10:40:09

Diagnostics Check	OK	10:34
Air Blank	0.000	10:34
Control Test	0.082	10:35
Air Blank	0.000	10:35
Subject Sample # 1	0.000	10:35
Breath Volume	2.140	10:35
Air Blank	0.000	10:36
Air Blank	0.000	10:38
Subject Sample # 2	0.000	10:38
Breath Volume	2.207	10:38
Air Blank	0.000	10:39
Control Test	0.082	10:39
Air Blank	0.000	10:40
Diagnostics Check	OK	10:40

Data Download Date/Time: 05/30/2019 15:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HICKS,WILLIAM,L
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: LOGAN,JONATHAN
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 03/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:39

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 04/04/2019 21:39:27

Diagnostics Check	OK	21:33
Air Blank	0.000	21:33
Control Test	0.081	21:34
Air Blank	0.000	21:34
Subject Sample # 1	0.231	21:35
Breath Volume	3.972	21:35
Air Blank	0.000	21:35
Air Blank	0.000	21:37
Subject Sample # 2	0.218	21:37
Breath Volume	3.832	21:37
Air Blank	0.000	21:38
Control Test	0.080	21:38
Air Blank	0.000	21:39
Diagnostics Check	OK	21:39

Data Download Date/Time: 04/23/2019 09:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LACEY,JOHN,J
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: LACEY,JOHN
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 03/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: LECOMPTE

DATE/TIME OF TEST: 04/07/2019 03:21:27

Diagnostics Check	OK	03:15
Air Blank	0.000	03:15
Control Test	0.081	03:15
Air Blank	0.000	03:16
Subject Sample # 1	0.207	03:17
Breath Volume	3.140	03:17
Air Blank	0.000	03:17
Air Blank	0.000	03:19
Subject Sample # 2	0.194	03:19
Breath Volume	2.222	03:19
Air Blank	0.000	03:20
Control Test	0.080	03:20
Air Blank	0.000	03:21
Diagnostics Check	OK	03:21

Data Download Date/Time: 04/23/2019 09:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OBYRNE,JAMES,W
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: OBYRNE,JAMES
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 03/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 11:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:01

SUBJECT'S LAST NAME: RADOMSKI

DATE/TIME OF TEST: 04/08/2019 11:33:37

Diagnostics Check	OK	11:27
Air Blank	0.000	11:27
Control Test	0.081	11:28
Air Blank	0.000	11:28
Subject Sample # 1	0.000	11:29
Breath Volume	1.308	11:29
Air Blank	0.000	11:29
Air Blank	0.000	11:31
Subject Sample # 2	0.000	11:32
Breath Volume	1.964	11:32
Air Blank	0.000	11:32
Control Test	0.080	11:33
Air Blank	0.000	11:33
Diagnostics Check	OK	11:33

Data Download Date/Time: 04/23/2019 09:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JASIUKIEWICZ,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 04/04/2019 19:07:51

Diagnostics Check	OK	19:00	
Air Blank	0.000	19:00	
Control Test	0.079	19:01	
Air Blank	0.000	19:01	
Subject Sample # 1	0.192	19:02	
Breath Volume	1.101	19:02	
Air Blank	0.000	19:02	
Air Blank	0.000	19:04	
Subject Sample # 2	0.191	19:06	
Breath Volume	1.296	19:06	
Air Blank	0.000	19:06	
Control Test	0.084	19:07	Interferent Detect
Air Blank	0.000	19:07	

Data Download Date/Time: 05/21/2019 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JASIUKIEWICZ,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 04/04/2019 19:16:22

Diagnostics Check	OK	19:09	
Air Blank	0.000	19:09	
Control Test	0.079	19:09	
Air Blank	0.000	19:10	
Subject Sample # 1	0.192	19:11	
Breath Volume	1.128	19:11	
Air Blank	0.000	19:12	
Air Blank	0.000	19:13	
Subject Sample # 2	0.186	19:14	
Breath Volume	1.347	19:14	
Air Blank	0.000	19:15	
Control Test	0.082	19:15	Interferent Detect
Air Blank	0.000	19:16	

Data Download Date/Time: 05/21/2019 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JASIUKIEWICZ,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 04/04/2019 19:23:42

Diagnostics Check	OK	19:17
Air Blank	0.000	19:18
Control Test	0.079	19:18
Air Blank	0.000	19:19
Subject Sample # 1	0.186	19:19
Breath Volume	1.210	19:19
Air Blank	0.000	19:20
Air Blank	0.000	19:21
Subject Sample # 2	0.180	19:22
Breath Volume	1.246	19:22
Air Blank	0.000	19:22
Control Test	0.080	19:23
Air Blank	0.000	19:23
Diagnostics Check	OK	19:23

Data Download Date/Time: 05/21/2019 10:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: ANCHUTKINA

DATE/TIME OF TEST: 04/05/2019 03:45:10

Diagnostics Check	OK	03:42	
Air Blank	0.000	03:42	
Control Test	0.079	03:43	
Air Blank	0.000	03:43	
Subject Sample	# 1	03:43	Subject Test Refused
Breath Volume	0.000	03:43	
Air Blank	0.000	03:44	
Control Test	0.079	03:44	
Air Blank	0.000	03:45	
Diagnostics Check	OK	03:45	

Data Download Date/Time: 05/21/2019 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:38

SUBJECT'S LAST NAME: REYNOLDS

DATE/TIME OF TEST: 04/05/2019 21:04:47

Diagnostics Check	OK	20:58
Air Blank	0.000	20:59
Control Test	0.079	20:59
Air Blank	0.000	21:00
Subject Sample # 1	0.000	21:00
Breath Volume	2.582	21:00
Air Blank	0.000	21:01
Air Blank	0.000	21:03
Subject Sample # 2	0.000	21:03
Breath Volume	1.789	21:03
Air Blank	0.000	21:03
Control Test	0.083	21:04
Air Blank	0.000	21:04
Diagnostics Check	OK	21:04

Data Download Date/Time: 05/21/2019 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: DOMINGUEZ

DATE/TIME OF TEST: 04/06/2019 03:02:14

Diagnostics Check	OK	02:52	
Air Blank	0.000	02:53	
Control Test	0.079	02:53	
Air Blank	0.000	02:54	
Subject Sample	# 1 0.178	02:57	Volume Not Met
Breath Volume	0.039	02:57	
Air Blank	0.000	02:57	
Air Blank	0.000	02:59	
Subject Sample	# 2	03:00	Subject Test Refused
Breath Volume	0.109	03:00	
Air Blank	0.000	03:01	
Control Test	0.079	03:01	
Air Blank	0.000	03:02	
Diagnostics Check	OK	03:02	

Data Download Date/Time: 05/21/2019 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: REPKO

DATE/TIME OF TEST: 04/07/2019 02:01:34

Diagnostics Check	OK	01:58	
Air Blank	0.000	01:59	
Control Test	0.079	01:59	
Air Blank	0.000	02:00	
Subject Sample	# 1	02:00	Subject Test Refused
Breath Volume	0.000	02:00	
Air Blank	0.000	02:00	
Control Test	0.078	02:00	
Air Blank	0.000	02:01	
Diagnostics Check	OK	02:01	

Data Download Date/Time: 05/21/2019 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUFF,ZACHARY,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: CZOP,BRYAN
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 04/09/2019 04:03:29

Diagnostics Check	OK	03:51
Air Blank	0.000	03:51
Control Test	0.077	03:52
Air Blank	0.000	03:52
Subject Sample # 1	0.205	03:53
Breath Volume	1.960	03:53
Air Blank	0.000	03:54
Air Blank	0.000	03:56
Subject Sample # 2	0.184	03:57
Breath Volume	1.308	03:57
Air Blank	0.000	03:58
Air Blank	0.000	03:59
Subject Sample # 3	0.203	04:01
Breath Volume	2.468	04:01
Air Blank	0.000	04:02
Control Test	0.083	04:02
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time: 05/21/2019 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUFF,ZACHARY,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: HUFF,ZACHARY
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: VENITSKY

DATE/TIME OF TEST: 04/10/2019 02:42:09

Diagnostics Check	OK	02:34
Air Blank	0.000	02:35
Control Test	0.079	02:35
Air Blank	0.000	02:36
Subject Sample # 1	0.000	02:37
Breath Volume	1.363	02:37
Air Blank	0.000	02:37
Air Blank	0.000	02:39
Subject Sample # 2	0.000	02:40
Breath Volume	1.320	02:40
Air Blank	0.000	02:41
Control Test	0.083	02:41
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 05/21/2019 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 04/11/2019 03:00:00

Diagnostics Check	OK	02:57	
Air Blank	0.000	02:57	
Control Test	0.078	02:58	
Air Blank	0.000	02:58	
Subject Sample	# 1	02:58	Subject Test Refused
Breath Volume	0.000	02:58	
Air Blank	0.000	02:59	
Control Test	0.079	02:59	
Air Blank	0.000	02:59	
Diagnostics Check	OK	03:00	

Data Download Date/Time: 05/21/2019 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: WRIGHT III

DATE/TIME OF TEST: 04/13/2019 04:42:50

Diagnostics Check	OK	04:35
Air Blank	0.000	04:36
Control Test	0.079	04:36
Air Blank	0.000	04:36
Subject Sample # 1	0.213	04:37
Breath Volume	1.121	04:37
Air Blank	0.000	04:37
Air Blank	0.000	04:39
Subject Sample # 2	0.198	04:41
Breath Volume	1.105	04:41
Air Blank	0.000	04:41
Control Test	0.083	04:42
Air Blank	0.000	04:42
Diagnostics Check	OK	04:42

Data Download Date/Time: 05/21/2019 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKINNER,SEAN,A
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: GOODMAN,JONATHAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:24

SUBJECT'S LAST NAME: MEEHAN

DATE/TIME OF TEST: 04/13/2019 05:48:16

Diagnostics Check	OK	05:44	
Air Blank	0.000	05:44	
Control Test	0.079	05:45	
Air Blank	0.000	05:45	
Subject Sample	# 1	05:46	Subject Test Refused
Breath Volume	0.000	05:46	
Air Blank	0.000	05:47	
Control Test	0.079	05:47	
Air Blank	0.000	05:48	
Diagnostics Check	OK	05:48	

Data Download Date/Time: 05/21/2019 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:46

SUBJECT'S LAST NAME: PAGEAU

DATE/TIME OF TEST: 04/13/2019 23:15:37

Diagnostics Check	OK	23:09
Air Blank	0.000	23:09
Control Test	0.079	23:10
Air Blank	0.000	23:10
Subject Sample # 1	0.250	23:11
Breath Volume	1.886	23:11
Air Blank	0.000	23:11
Air Blank	0.000	23:13
Subject Sample # 2	0.244	23:14
Breath Volume	1.621	23:14
Air Blank	0.000	23:14
Control Test	0.081	23:15
Air Blank	0.000	23:15
Diagnostics Check	OK	23:15

Data Download Date/Time: 05/21/2019 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: GILCHRIST

DATE/TIME OF TEST: 04/17/2019 21:41:31

Diagnostics Check	OK	21:35	
Air Blank	0.000	21:35	
Control Test	0.079	21:36	
Air Blank	0.000	21:36	
Subject Sample # 1	0.034	21:37	
Breath Volume	2.691	21:37	
Air Blank	0.000	21:37	
Air Blank	0.000	21:39	
Subject Sample # 2	0.033	21:40	
Breath Volume	2.316	21:40	
Air Blank	0.000	21:40	
Control Test	0.086	21:40	Control Outside Tolerance
Air Blank	0.000	21:41	

Data Download Date/Time: 05/21/2019 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: LEENEY

DATE/TIME OF TEST: 04/18/2019 00:55:16

Diagnostics Check	OK	00:49
Air Blank	0.000	00:49
Control Test	0.079	00:49
Air Blank	0.000	00:50
Subject Sample # 1	0.071	00:50
Breath Volume	1.300	00:50
Air Blank	0.000	00:51
Air Blank	0.000	00:53
Subject Sample # 2	0.075	00:53
Breath Volume	1.808	00:53
Air Blank	0.000	00:54
Control Test	0.082	00:54
Air Blank	0.000	00:55
Diagnostics Check	OK	00:55

Data Download Date/Time: 05/21/2019 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: ESKELES

DATE/TIME OF TEST: 04/18/2019 21:28:16

Diagnostics Check	OK	21:25	
Air Blank	0.000	21:25	
Control Test	0.079	21:26	
Air Blank	0.000	21:26	
Subject Sample	# 1	21:26	Subject Test Refused
Breath Volume	0.000	21:26	
Air Blank	0.000	21:27	
Control Test	0.078	21:27	
Air Blank	0.000	21:28	
Diagnostics Check	OK	21:28	

Data Download Date/Time: 05/21/2019 10:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: CAMACHO

DATE/TIME OF TEST: 04/19/2019 03:46:01

Diagnostics Check	OK	03:43	
Air Blank	0.000	03:43	
Control Test	0.079	03:43	
Air Blank	0.000	03:44	
Subject Sample	# 1	03:44	Subject Test Refused
Breath Volume	0.000	03:44	
Air Blank	0.000	03:45	
Control Test	0.079	03:45	
Air Blank	0.000	03:45	
Diagnostics Check	OK	03:46	

Data Download Date/Time: 05/21/2019 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOYLE,THOMAS,E
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: BOYLE,THOMAS
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: PAYNE

DATE/TIME OF TEST: 04/19/2019 04:09:25

Diagnostics Check	OK	04:01	
Air Blank	0.000	04:01	
Control Test	0.079	04:02	
Air Blank	0.000	04:02	
Subject Sample # 1	0.162	04:02	
Breath Volume	1.226	04:02	
Air Blank	0.000	04:03	
Air Blank	0.000	04:05	
Subject Sample # 2	0.157	04:07	
Breath Volume	1.832	04:07	
Air Blank	0.000	04:08	
Control Test	0.085	04:08	Interferent Detect
Air Blank	0.000	04:09	

Data Download Date/Time: 05/21/2019 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOYLE,THOMAS,E
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: BOYLE,THOMAS
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: PAYNE

DATE/TIME OF TEST: 04/19/2019 04:23:42

Diagnostics Check	OK	04:15
Air Blank	0.000	04:16
Control Test	0.079	04:16
Air Blank	0.000	04:16
Subject Sample # 1	0.148	04:19
Breath Volume	1.710	04:19
Air Blank	0.000	04:20
Air Blank	0.000	04:21
Subject Sample # 2	0.157	04:22
Breath Volume	1.843	04:22
Air Blank	0.000	04:22
Control Test	0.081	04:23
Air Blank	0.000	04:23
Diagnostics Check	OK	04:23

Data Download Date/Time: 05/21/2019 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 04/20/2019 21:48:30

Diagnostics Check	OK	21:42
Air Blank	0.000	21:42
Control Test	0.079	21:43
Air Blank	0.000	21:43
Subject Sample # 1	0.147	21:44
Breath Volume	1.167	21:44
Air Blank	0.000	21:44
Air Blank	0.000	21:46
Subject Sample # 2	0.152	21:47
Breath Volume	1.359	21:47
Air Blank	0.000	21:47
Control Test	0.081	21:47
Air Blank	0.000	21:48
Diagnostics Check	OK	21:48

Data Download Date/Time: 05/21/2019 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:23

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 04/25/2019 18:54:40

Diagnostics Check	OK	18:48
Air Blank	0.000	18:48
Control Test	0.080	18:49
Air Blank	0.000	18:49
Subject Sample # 1	0.125	18:50
Breath Volume	1.304	18:50
Air Blank	0.000	18:50
Air Blank	0.000	18:52
Subject Sample # 2	0.119	18:53
Breath Volume	1.105	18:53
Air Blank	0.000	18:53
Control Test	0.081	18:54
Air Blank	0.000	18:54
Diagnostics Check	OK	18:54

Data Download Date/Time: 05/21/2019 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: CONNICK

DATE/TIME OF TEST: 04/26/2019 23:11:17

Diagnostics Check	OK	23:00	
Air Blank	0.000	23:00	
Control Test	0.079	23:01	
Air Blank	0.000	23:01	
Subject Sample	# 1	23:04	No Sample Provided
Breath Volume	0.000	23:04	
Air Blank	0.000	23:05	
Air Blank	0.000	23:07	
Subject Sample	# 2	23:09	Subject Test Refused
Breath Volume	0.000	23:09	
Air Blank	0.000	23:10	
Control Test	0.079	23:10	
Air Blank	0.000	23:11	
Diagnostics Check	OK	23:11	

Data Download Date/Time: 05/21/2019 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: ALDER

DATE/TIME OF TEST: 04/28/2019 00:31:49

Diagnostics Check	OK	00:25
Air Blank	0.000	00:26
Control Test	0.079	00:26
Air Blank	0.000	00:26
Subject Sample # 1	0.081	00:27
Breath Volume	1.210	00:27
Air Blank	0.000	00:28
Air Blank	0.000	00:29
Subject Sample # 2	0.079	00:30
Breath Volume	1.339	00:30
Air Blank	0.000	00:30
Control Test	0.080	00:31
Air Blank	0.000	00:31
Diagnostics Check	OK	00:31

Data Download Date/Time: 05/21/2019 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOYLE, THOMAS, E
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: DIAZ, LUIS
ARRESTING OFFICER AGENCY: Kenneth City PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: KRAMER

DATE/TIME OF TEST: 04/28/2019 23:57:08

Diagnostics Check	OK	23:50
Air Blank	0.000	23:51
Control Test	0.078	23:51
Air Blank	0.000	23:52
Subject Sample # 1	0.179	23:52
Breath Volume	1.113	23:52
Air Blank	0.000	23:53
Air Blank	0.000	23:55
Subject Sample # 2	0.166	23:55
Breath Volume	1.246	23:55
Air Blank	0.000	23:56
Control Test	0.082	23:56
Air Blank	0.000	23:57
Diagnostics Check	OK	23:57

Data Download Date/Time: 05/21/2019 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKINNER,SEAN,A
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: RAYMOND,CHRISTIAN
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: CRIDER

DATE/TIME OF TEST: 04/29/2019 01:38:52

Diagnostics Check	OK	01:35	
Air Blank	0.000	01:36	
Control Test	0.078	01:36	
Air Blank	0.000	01:36	
Subject Sample	# 1	01:37	Subject Test Refused
Breath Volume	0.000	01:37	
Air Blank	0.000	01:37	
Control Test	0.079	01:38	
Air Blank	0.000	01:38	
Diagnostics Check	OK	01:38	

Data Download Date/Time: 05/21/2019 10:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOYLE,THOMAS,E
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: BOYLE,THOMAS
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: CONKLIN III

DATE/TIME OF TEST: 04/29/2019 04:27:43

Diagnostics Check	OK	04:24	
Air Blank	0.000	04:25	
Control Test	0.079	04:25	
Air Blank	0.000	04:26	
Subject Sample	# 1	04:26	Subject Test Refused
Breath Volume	0.000	04:26	
Air Blank	0.000	04:26	
Control Test	0.079	04:27	
Air Blank	0.000	04:27	
Diagnostics Check	OK	04:27	

Data Download Date/Time: 05/21/2019 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: COCONUT CREEK PD
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: COCONUT CREEK PD
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: ARMAND

DATE/TIME OF TEST: 04/12/2019 00:56:50

Diagnostics Check	OK	00:50
Air Blank	0.000	00:51
Control Test	0.079	00:51
Air Blank	0.000	00:51
Subject Sample # 1	0.083	00:52
Breath Volume	1.820	00:52
Air Blank	0.000	00:53
Air Blank	0.000	00:55
Subject Sample # 2	0.079	00:55
Breath Volume	1.171	00:55
Air Blank	0.000	00:55
Control Test	0.078	00:56
Air Blank	0.000	00:56
Diagnostics Check	OK	00:56

Data Download Date/Time: 04/29/2019 15:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: COCONUT CREEK PD
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: COCONUT CREEK PD
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: SUFFRA

DATE/TIME OF TEST: 04/24/2019 23:24:23

Diagnostics Check	OK	23:21	
Air Blank	0.000	23:22	
Control Test	0.078	23:22	
Air Blank	0.000	23:22	
Subject Sample	# 1	23:23	Subject Test Refused
Breath Volume	0.000	23:23	
Air Blank	0.000	23:23	
Control Test	0.078	23:23	
Air Blank	0.000	23:24	
Diagnostics Check	OK	23:24	

Data Download Date/Time: 04/29/2019 15:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: COCONUT CREEK PD
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: COCONUT CREEK PD
LAST AGENCY INSPECTION: 04/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: ZABRISKIE

DATE/TIME OF TEST: 04/29/2019 22:57:13

Diagnostics Check	OK	22:54	
Air Blank	0.000	22:54	
Control Test	0.078	22:55	
Air Blank	0.000	22:55	
Subject Sample	# 1	22:55	Subject Test Refused
Breath Volume	0.000	22:55	
Air Blank	0.000	22:56	
Control Test	0.079	22:56	
Air Blank	0.000	22:57	
Diagnostics Check	OK	22:57	

Data Download Date/Time: 05/29/2019 15:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: PALMER

DATE/TIME OF TEST: 04/01/2019 00:09:03

Diagnostics Check	OK	00:00	
Air Blank	0.000	00:01	
Control Test	0.080	00:01	
Air Blank	0.000	00:01	
Subject Sample	# 1 0.110	00:05	Volume Not Met
Breath Volume	0.113	00:05	
Air Blank	0.000	00:05	
Air Blank	0.000	00:07	
Subject Sample	# 2	00:07	Subject Test Refused
Breath Volume	0.000	00:07	
Air Blank	0.000	00:08	
Control Test	0.078	00:08	
Air Blank	0.000	00:08	
Diagnostics Check	OK	00:09	

Data Download Date/Time: 04/16/2019 15:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SORRELLS,M
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: HUTTO

DATE/TIME OF TEST: 04/01/2019 03:40:24

Diagnostics Check	OK	03:34
Air Blank	0.000	03:34
Control Test	0.080	03:34
Air Blank	0.000	03:35
Subject Sample # 1	0.093	03:35
Breath Volume	3.121	03:35
Air Blank	0.000	03:36
Air Blank	0.000	03:38
Subject Sample # 2	0.091	03:38
Breath Volume	3.058	03:38
Air Blank	0.000	03:39
Control Test	0.079	03:39
Air Blank	0.000	03:40
Diagnostics Check	OK	03:40

Data Download Date/Time: 04/16/2019 15:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MCFARLAND,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:22

SUBJECT'S LAST NAME: HUNTER

DATE/TIME OF TEST: 04/01/2019 05:51:50

Diagnostics Check	OK	05:42
Air Blank	0.000	05:42
Control Test	0.080	05:43
Air Blank	0.000	05:43
Subject Sample # 1	0.172	05:45
Breath Volume	2.734	05:45
Air Blank	0.000	05:45
Air Blank	0.000	05:47
Subject Sample # 2	0.166	05:50
Breath Volume	1.132	05:50
Air Blank	0.000	05:50
Control Test	0.079	05:51
Air Blank	0.000	05:51
Diagnostics Check	OK	05:51

Data Download Date/Time: 04/16/2019 15:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MORGAN,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: TUTT

DATE/TIME OF TEST: 04/02/2019 02:01:41

Diagnostics Check	OK	01:55
Air Blank	0.000	01:55
Control Test	0.080	01:55
Air Blank	0.000	01:56
Subject Sample # 1	0.146	01:56
Breath Volume	1.765	01:56
Air Blank	0.000	01:57
Air Blank	0.000	01:59
Subject Sample # 2	0.156	02:00
Breath Volume	2.687	02:00
Air Blank	0.000	02:00
Control Test	0.079	02:01
Air Blank	0.000	02:01
Diagnostics Check	OK	02:01

Data Download Date/Time: 04/16/2019 15:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: GOLDEN

DATE/TIME OF TEST: 04/03/2019 02:15:04

Diagnostics Check	OK	02:09
Air Blank	0.000	02:09
Control Test	0.080	02:09
Air Blank	0.000	02:10
Subject Sample # 1	0.000	02:10
Breath Volume	2.316	02:10
Air Blank	0.000	02:11
Air Blank	0.000	02:13
Subject Sample # 2	0.000	02:13
Breath Volume	1.871	02:13
Air Blank	0.000	02:14
Control Test	0.080	02:14
Air Blank	0.000	02:14
Diagnostics Check	OK	02:15

Data Download Date/Time: 04/16/2019 15:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS, CHRISTOPHER, D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNS, B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: BRAZIEL

DATE/TIME OF TEST: 04/03/2019 04:11:07

Diagnostics Check	OK	04:04
Air Blank	0.000	04:05
Control Test	0.080	04:05
Air Blank	0.000	04:06
Subject Sample # 1	0.145	04:06
Breath Volume	2.117	04:06
Air Blank	0.000	04:07
Air Blank	0.000	04:09
Subject Sample # 2	0.147	04:09
Breath Volume	2.234	04:09
Air Blank	0.000	04:10
Control Test	0.079	04:10
Air Blank	0.000	04:11
Diagnostics Check	OK	04:11

Data Download Date/Time: 04/16/2019 15:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: SETH

DATE/TIME OF TEST: 04/04/2019 01:13:26

Diagnostics Check	OK	01:10	
Air Blank	0.000	01:11	
Control Test	0.080	01:11	
Air Blank	0.000	01:11	
Subject Sample	# 1	01:12	Subject Test Refused
Breath Volume	0.000	01:12	
Air Blank	0.000	01:12	
Control Test	0.079	01:12	
Air Blank	0.000	01:13	
Diagnostics Check	OK	01:13	

Data Download Date/Time: 04/16/2019 15:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS, CHRISTOPHER, D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY, T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: RAMSEY

DATE/TIME OF TEST: 04/04/2019 01:55:23

Diagnostics Check	OK	01:48
Air Blank	0.000	01:48
Control Test	0.080	01:48
Air Blank	0.000	01:49
Subject Sample # 1	0.142	01:50
Breath Volume	1.519	01:50
Air Blank	0.000	01:50
Air Blank	0.000	01:52
Subject Sample # 2	0.142	01:53
Breath Volume	1.976	01:53
Air Blank	0.000	01:54
Control Test	0.078	01:54
Air Blank	0.000	01:55
Diagnostics Check	OK	01:55

Data Download Date/Time: 04/16/2019 15:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VEALE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: FRAZIER

DATE/TIME OF TEST: 04/04/2019 04:04:30

Diagnostics Check	OK	03:58
Air Blank	0.000	03:58
Control Test	0.079	03:59
Air Blank	0.000	03:59
Subject Sample # 1	0.000	04:00
Breath Volume	3.632	04:00
Air Blank	0.000	04:00
Air Blank	0.000	04:02
Subject Sample # 2	0.000	04:03
Breath Volume	3.644	04:03
Air Blank	0.000	04:03
Control Test	0.079	04:03
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

Data Download Date/Time: 04/16/2019 15:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HHYSMITH,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:33

SUBJECT'S LAST NAME: ARMIJO

DATE/TIME OF TEST: 04/05/2019 05:56:30

Diagnostics Check	OK	05:53	
Air Blank	0.000	05:54	
Control Test	0.080	05:54	
Air Blank	0.000	05:54	
Subject Sample	# 1	05:55	Subject Test Refused
Breath Volume	0.000	05:55	
Air Blank	0.000	05:55	
Control Test	0.079	05:55	
Air Blank	0.000	05:56	
Diagnostics Check	OK	05:56	

Data Download Date/Time: 04/16/2019 15:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 04/05/2019 23:58:17

Diagnostics Check	OK	23:48
Air Blank	0.000	23:49
Control Test	0.080	23:49
Air Blank	0.000	23:49
Subject Sample # 1	0.266	23:52
Breath Volume	1.597	23:52
Air Blank	0.000	23:53
Air Blank	0.000	23:54
Subject Sample # 2	0.259	23:56
Breath Volume	1.746	23:56
Air Blank	0.000	23:57
Control Test	0.077	23:57
Air Blank	0.000	23:58
Diagnostics Check	OK	23:58

Data Download Date/Time: 04/16/2019 15:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: GRANILLO

DATE/TIME OF TEST: 04/06/2019 03:15:01

Diagnostics Check	OK	03:08
Air Blank	0.000	03:08
Control Test	0.079	03:09
Air Blank	0.000	03:09
Subject Sample # 1	0.218	03:10
Breath Volume	3.023	03:10
Air Blank	0.000	03:11
Air Blank	0.000	03:12
Subject Sample # 2	0.210	03:13
Breath Volume	2.246	03:13
Air Blank	0.000	03:14
Control Test	0.078	03:14
Air Blank	0.000	03:14
Diagnostics Check	OK	03:15

Data Download Date/Time: 04/16/2019 15:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARREN, SAVANNAH, B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DAVIS, R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 12:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:26

SUBJECT'S LAST NAME: WHITLOW

DATE/TIME OF TEST: 04/06/2019 15:55:02

Diagnostics Check	OK	15:46
Air Blank	0.000	15:47
Control Test	0.078	15:47
Air Blank	0.000	15:47
Subject Sample # 1	0.296	15:50
Breath Volume	2.773	15:50
Air Blank	0.000	15:51
Air Blank	0.000	15:52
Subject Sample # 2	0.301	15:53
Breath Volume	3.109	15:53
Air Blank	0.000	15:54
Control Test	0.076	15:54
Air Blank	0.000	15:54
Diagnostics Check	OK	15:55

Data Download Date/Time: 04/16/2019 15:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARREN, SAVANNAH, B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES, J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: SPARKS

DATE/TIME OF TEST: 04/06/2019 18:38:25

Diagnostics Check	OK	18:31
Air Blank	0.000	18:31
Control Test	0.079	18:31
Air Blank	0.000	18:32
Subject Sample # 1	0.186	18:33
Breath Volume	1.765	18:33
Air Blank	0.000	18:34
Air Blank	0.000	18:36
Subject Sample # 2	0.195	18:36
Breath Volume	1.640	18:36
Air Blank	0.000	18:37
Control Test	0.078	18:37
Air Blank	0.000	18:38
Diagnostics Check	OK	18:38

Data Download Date/Time: 04/16/2019 15:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARREN, SAVANNAH, B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ, J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:16

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 04/07/2019 20:43:27

Diagnostics Check	OK	20:36
Air Blank	0.000	20:37
Control Test	0.079	20:37
Air Blank	0.000	20:38
Subject Sample # 1	0.221	20:38
Breath Volume	3.414	20:38
Air Blank	0.000	20:39
Air Blank	0.000	20:41
Subject Sample # 2	0.234	20:41
Breath Volume	3.449	20:41
Air Blank	0.000	20:42
Control Test	0.078	20:42
Air Blank	0.000	20:43
Diagnostics Check	OK	20:43

Data Download Date/Time: 04/16/2019 15:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARREN, SAVANNAH, B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOORE, B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: FOURES

DATE/TIME OF TEST: 04/07/2019 21:28:27

Diagnostics Check	OK	21:25	
Air Blank	0.000	21:26	
Control Test	0.079	21:26	
Air Blank	0.000	21:27	
Subject Sample	# 1	21:27	Sequence Aborted
Breath Volume	0.000	21:27	
Air Blank	0.000	21:28	

Data Download Date/Time: 04/16/2019 15:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARREN, SAVANNAH, B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOORE, B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: FOURES

DATE/TIME OF TEST: 04/07/2019 21:47:42

Diagnostics Check	OK	21:41
Air Blank	0.000	21:41
Control Test	0.079	21:41
Air Blank	0.000	21:42
Subject Sample # 1	0.100	21:43
Breath Volume	2.804	21:43
Air Blank	0.000	21:43
Air Blank	0.000	21:45
Subject Sample # 2	0.097	21:46
Breath Volume	2.511	21:46
Air Blank	0.000	21:46
Control Test	0.079	21:47
Air Blank	0.000	21:47
Diagnostics Check	OK	21:47

Data Download Date/Time: 04/16/2019 15:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMONS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: DUNCAN

DATE/TIME OF TEST: 04/07/2019 23:14:07

Diagnostics Check	OK	23:04
Air Blank	0.000	23:05
Control Test	0.079	23:05
Air Blank	0.000	23:06
Subject Sample # 1	0.166	23:08
Breath Volume	1.234	23:08
Air Blank	0.000	23:09
Air Blank	0.000	23:10
Subject Sample # 2	0.160	23:12
Breath Volume	1.441	23:12
Air Blank	0.000	23:13
Control Test	0.077	23:13
Air Blank	0.000	23:14
Diagnostics Check	OK	23:14

Data Download Date/Time: 04/16/2019 15:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: GLOVER

DATE/TIME OF TEST: 04/07/2019 23:50:25

Diagnostics Check	OK	23:44
Air Blank	0.000	23:44
Control Test	0.079	23:45
Air Blank	0.000	23:45
Subject Sample # 1	0.133	23:46
Breath Volume	1.195	23:46
Air Blank	0.000	23:46
Air Blank	0.000	23:48
Subject Sample # 2	0.136	23:48
Breath Volume	2.261	23:48
Air Blank	0.000	23:49
Control Test	0.079	23:49
Air Blank	0.000	23:50
Diagnostics Check	OK	23:50

Data Download Date/Time: 04/16/2019 15:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: WELLS

DATE/TIME OF TEST: 04/08/2019 00:33:50

Diagnostics Check	OK	00:27
Air Blank	0.000	00:28
Control Test	0.078	00:28
Air Blank	0.000	00:28
Subject Sample # 1	0.090	00:29
Breath Volume	2.105	00:29
Air Blank	0.000	00:29
Air Blank	0.000	00:31
Subject Sample # 2	0.093	00:32
Breath Volume	2.179	00:32
Air Blank	0.000	00:32
Control Test	0.079	00:33
Air Blank	0.000	00:33
Diagnostics Check	OK	00:33

Data Download Date/Time: 04/16/2019 15:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: FORTUNE

DATE/TIME OF TEST: 04/08/2019 03:51:55

Diagnostics Check	OK	03:45
Air Blank	0.000	03:46
Control Test	0.079	03:46
Air Blank	0.000	03:46
Subject Sample # 1	0.201	03:47
Breath Volume	1.949	03:47
Air Blank	0.000	03:48
Air Blank	0.000	03:49
Subject Sample # 2	0.207	03:50
Breath Volume	1.269	03:50
Air Blank	0.000	03:50
Control Test	0.079	03:51
Air Blank	0.000	03:51
Diagnostics Check	OK	03:51

Data Download Date/Time: 04/16/2019 15:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HARPIN,R
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:19

SUBJECT'S LAST NAME: GROTE

DATE/TIME OF TEST: 04/08/2019 05:46:25

Diagnostics Check	OK	05:39
Air Blank	0.000	05:40
Control Test	0.080	05:40
Air Blank	0.000	05:41
Subject Sample # 1	0.176	05:41
Breath Volume	1.511	05:41
Air Blank	0.000	05:42
Air Blank	0.000	05:44
Subject Sample # 2	0.172	05:44
Breath Volume	1.902	05:44
Air Blank	0.000	05:45
Control Test	0.079	05:45
Air Blank	0.000	05:46
Diagnostics Check	OK	05:46

Data Download Date/Time: 04/16/2019 15:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILL,D,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:58

SUBJECT'S LAST NAME: HOLLOMAN

DATE/TIME OF TEST: 04/08/2019 20:25:37

Diagnostics Check	OK	20:18
Air Blank	0.000	20:19
Control Test	0.080	20:19
Air Blank	0.000	20:20
Subject Sample # 1	0.126	20:20
Breath Volume	1.828	20:20
Air Blank	0.000	20:21
Air Blank	0.000	20:23
Subject Sample # 2	0.125	20:24
Breath Volume	2.035	20:24
Air Blank	0.000	20:24
Control Test	0.080	20:25
Air Blank	0.000	20:25
Diagnostics Check	OK	20:25

Data Download Date/Time: 04/16/2019 15:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILL,D,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:23

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 04/08/2019 22:50:19

Diagnostics Check	OK	22:43
Air Blank	0.000	22:44
Control Test	0.080	22:44
Air Blank	0.000	22:45
Subject Sample # 1	0.256	22:45
Breath Volume	1.410	22:45
Air Blank	0.000	22:46
Air Blank	0.000	22:48
Subject Sample # 2	0.244	22:48
Breath Volume	1.308	22:48
Air Blank	0.000	22:49
Control Test	0.079	22:49
Air Blank	0.000	22:50
Diagnostics Check	OK	22:50

Data Download Date/Time: 04/16/2019 15:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VANHOF,J
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: STRICKLAND

DATE/TIME OF TEST: 04/09/2019 03:04:21

Diagnostics Check	OK	02:51	
Air Blank	0.000	02:51	
Control Test	0.080	02:51	
Air Blank	0.000	02:52	
Subject Sample	# 1 0.129	02:53	
Breath Volume	1.160	02:53	
Air Blank	0.000	02:54	
Air Blank	0.000	02:56	
Subject Sample	# 2 0.121	02:59	Volume Not Met
Breath Volume	0.625	02:59	
Air Blank	0.000	03:00	
Air Blank	0.000	03:02	
Subject Sample	# 3 0.131	03:02	
Breath Volume	1.343	03:02	
Air Blank	0.000	03:03	
Control Test	0.077	03:03	
Air Blank	0.000	03:04	
Diagnostics Check	OK	03:04	

Data Download Date/Time: 04/16/2019 15:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 04/09/2019 04:13:51

Diagnostics Check	OK	04:07
Air Blank	0.000	04:07
Control Test	0.079	04:08
Air Blank	0.000	04:08
Subject Sample # 1	0.185	04:09
Breath Volume	2.117	04:09
Air Blank	0.000	04:09
Air Blank	0.000	04:11
Subject Sample # 2	0.189	04:12
Breath Volume	3.312	04:12
Air Blank	0.000	04:12
Control Test	0.079	04:13
Air Blank	0.000	04:13
Diagnostics Check	OK	04:13

Data Download Date/Time: 04/16/2019 15:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: BOYKIN

DATE/TIME OF TEST: 04/10/2019 00:06:09

Diagnostics Check	OK	00:03	
Air Blank	0.000	00:03	
Control Test	0.080	00:04	
Air Blank	0.000	00:04	
Subject Sample	# 1	00:04	Subject Test Refused
Breath Volume	0.000	00:04	
Air Blank	0.000	00:05	
Control Test	0.079	00:05	
Air Blank	0.000	00:06	
Diagnostics Check	OK	00:06	

Data Download Date/Time: 04/16/2019 15:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JONES,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: GLENN

DATE/TIME OF TEST: 04/10/2019 02:19:32

Diagnostics Check	OK	02:16	
Air Blank	0.000	02:17	
Control Test	0.080	02:17	
Air Blank	0.000	02:17	
Subject Sample	# 1	02:18	Subject Test Refused
Breath Volume	0.000	02:18	
Air Blank	0.000	02:18	
Control Test	0.079	02:18	
Air Blank	0.000	02:19	
Diagnostics Check	OK	02:19	

Data Download Date/Time: 04/16/2019 15:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS, CHRISTOPHER, Y
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ, M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: WITT

DATE/TIME OF TEST: 04/11/2019 02:45:55

Diagnostics Check	OK	02:38
Air Blank	0.000	02:38
Control Test	0.080	02:39
Air Blank	0.000	02:39
Subject Sample # 1	0.184	02:40
Breath Volume	1.570	02:40
Air Blank	0.000	02:40
Air Blank	0.000	02:42
Subject Sample # 2	0.172	02:44
Breath Volume	1.378	02:44
Air Blank	0.000	02:44
Control Test	0.079	02:45
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 04/16/2019 15:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DEAN,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 04/11/2019 05:46:44

Diagnostics Check	OK	05:40
Air Blank	0.000	05:40
Control Test	0.079	05:41
Air Blank	0.000	05:41
Subject Sample # 1	0.136	05:42
Breath Volume	2.058	05:42
Air Blank	0.000	05:42
Air Blank	0.000	05:44
Subject Sample # 2	0.133	05:45
Breath Volume	2.000	05:45
Air Blank	0.000	05:45
Control Test	0.080	05:46
Air Blank	0.000	05:46
Diagnostics Check	OK	05:46

Data Download Date/Time: 04/16/2019 15:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS, CHRISTOPHER, D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY, T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 04/12/2019 02:18:14

Diagnostics Check	OK	02:11
Air Blank	0.000	02:12
Control Test	0.080	02:12
Air Blank	0.000	02:13
Subject Sample # 1	0.122	02:13
Breath Volume	2.328	02:13
Air Blank	0.000	02:14
Air Blank	0.000	02:16
Subject Sample # 2	0.114	02:16
Breath Volume	2.253	02:16
Air Blank	0.000	02:17
Control Test	0.080	02:17
Air Blank	0.000	02:18
Diagnostics Check	OK	02:18

Data Download Date/Time: 04/16/2019 15:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS, CHRISTOPHER, D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HARD, W
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:06

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 04/12/2019 05:33:20

Diagnostics Check	OK	05:26
Air Blank	0.000	05:27
Control Test	0.080	05:27
Air Blank	0.000	05:28
Subject Sample # 1	0.213	05:28
Breath Volume	2.558	05:28
Air Blank	0.000	05:29
Air Blank	0.000	05:31
Subject Sample # 2	0.210	05:31
Breath Volume	2.359	05:31
Air Blank	0.000	05:32
Control Test	0.079	05:32
Air Blank	0.000	05:33
Diagnostics Check	OK	05:33

Data Download Date/Time: 04/16/2019 15:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: LASTER

DATE/TIME OF TEST: 04/14/2019 00:50:38

Diagnostics Check	OK	00:44
Air Blank	0.000	00:44
Control Test	0.080	00:44
Air Blank	0.000	00:45
Subject Sample # 1	0.158	00:46
Breath Volume	2.699	00:46
Air Blank	0.000	00:46
Air Blank	0.000	00:48
Subject Sample # 2	0.160	00:49
Breath Volume	1.503	00:49
Air Blank	0.000	00:49
Control Test	0.079	00:50
Air Blank	0.000	00:50
Diagnostics Check	OK	00:50

Data Download Date/Time: 04/16/2019 15:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNSON,G
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: POWELL

DATE/TIME OF TEST: 04/14/2019 05:09:37

Diagnostics Check	OK	04:58
Air Blank	0.000	04:58
Control Test	0.080	04:59
Air Blank	0.000	04:59
Subject Sample # 1	0.178	05:00
Breath Volume	1.890	05:00
Air Blank	0.000	05:01
Air Blank	0.000	05:03
Subject Sample # 2	0.155	05:04
Breath Volume	1.902	05:04
Air Blank	0.000	05:05
Air Blank	0.000	05:06
Subject Sample # 3	0.159	05:07
Breath Volume	1.410	05:07
Air Blank	0.000	05:08
Control Test	0.078	05:08
Air Blank	0.000	05:09
Diagnostics Check	OK	05:09

Data Download Date/Time: 04/16/2019 15:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS, CHRISTOPHER, D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TORRES, W
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 04/15/2019 04:09:24

Diagnostics Check	OK	04:02
Air Blank	0.000	04:02
Control Test	0.080	04:03
Air Blank	0.000	04:03
Subject Sample # 1	0.180	04:04
Breath Volume	2.003	04:04
Air Blank	0.000	04:05
Air Blank	0.000	04:07
Subject Sample # 2	0.173	04:07
Breath Volume	1.281	04:07
Air Blank	0.000	04:08
Control Test	0.079	04:08
Air Blank	0.000	04:09
Diagnostics Check	OK	04:09

Data Download Date/Time: 04/16/2019 15:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FILBERT,E
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:33

SUBJECT'S LAST NAME: HOBBS

DATE/TIME OF TEST: 04/16/2019 04:56:40

Diagnostics Check	OK	04:53	
Air Blank	0.000	04:54	
Control Test	0.080	04:54	
Air Blank	0.000	04:55	
Subject Sample	# 1	04:55	Subject Test Refused
Breath Volume	0.000	04:55	
Air Blank	0.000	04:55	
Control Test	0.080	04:56	
Air Blank	0.000	04:56	
Diagnostics Check	OK	04:56	

Data Download Date/Time: 04/16/2019 15:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS, CHRISTOPHER, D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNSON, G
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: O'BRIEN

DATE/TIME OF TEST: 04/18/2019 00:55:53

Diagnostics Check	OK	00:49
Air Blank	0.000	00:49
Control Test	0.080	00:49
Air Blank	0.000	00:50
Subject Sample # 1	0.136	00:51
Breath Volume	2.773	00:51
Air Blank	0.000	00:51
Air Blank	0.000	00:53
Subject Sample # 2	0.141	00:54
Breath Volume	2.289	00:54
Air Blank	0.000	00:54
Control Test	0.079	00:55
Air Blank	0.000	00:55
Diagnostics Check	OK	00:55

Data Download Date/Time: 05/01/2019 12:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS, CHRISTOPHER, D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE, J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: HOOD

DATE/TIME OF TEST: 04/18/2019 02:18:31

Diagnostics Check	OK	02:12
Air Blank	0.000	02:12
Control Test	0.080	02:13
Air Blank	0.000	02:13
Subject Sample # 1	0.095	02:14
Breath Volume	1.789	02:14
Air Blank	0.000	02:14
Air Blank	0.000	02:16
Subject Sample # 2	0.090	02:16
Breath Volume	1.187	02:16
Air Blank	0.000	02:17
Control Test	0.079	02:17
Air Blank	0.000	02:18
Diagnostics Check	OK	02:18

Data Download Date/Time: 05/01/2019 12:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: HEFFNER

DATE/TIME OF TEST: 04/18/2019 21:06:44

Diagnostics Check	OK	21:00
Air Blank	0.000	21:00
Control Test	0.080	21:01
Air Blank	0.000	21:01
Subject Sample # 1	0.193	21:02
Breath Volume	2.839	21:02
Air Blank	0.000	21:02
Air Blank	0.000	21:04
Subject Sample # 2	0.188	21:05
Breath Volume	2.718	21:05
Air Blank	0.000	21:05
Control Test	0.079	21:06
Air Blank	0.000	21:06
Diagnostics Check	OK	21:06

Data Download Date/Time: 05/01/2019 12:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DIAZ,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: MELCOLM

DATE/TIME OF TEST: 04/20/2019 03:03:16

Diagnostics Check	OK	03:00	
Air Blank	0.000	03:00	
Control Test	0.080	03:01	
Air Blank	0.000	03:01	
Subject Sample	# 1	03:01	Subject Test Refused
Breath Volume	0.000	03:01	
Air Blank	0.000	03:02	
Control Test	0.080	03:02	
Air Blank	0.000	03:03	
Diagnostics Check	OK	03:03	

Data Download Date/Time: 05/01/2019 12:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ULYSSES,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: MIMS

DATE/TIME OF TEST: 04/20/2019 05:55:04

Diagnostics Check	OK	05:52	
Air Blank	0.000	05:52	
Control Test	0.081	05:53	
Air Blank	0.000	05:53	
Subject Sample	# 1	05:53	Subject Test Refused
Breath Volume	0.000	05:53	
Air Blank	0.000	05:54	
Control Test	0.080	05:54	
Air Blank	0.000	05:54	
Diagnostics Check	OK	05:55	

Data Download Date/Time: 05/01/2019 12:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: WHIPPLE

DATE/TIME OF TEST: 04/20/2019 21:16:53

Diagnostics Check	OK	21:10
Air Blank	0.000	21:10
Control Test	0.080	21:11
Air Blank	0.000	21:11
Subject Sample # 1	0.000	21:12
Breath Volume	3.375	21:12
Air Blank	0.000	21:12
Air Blank	0.000	21:14
Subject Sample # 2	0.000	21:15
Breath Volume	3.167	21:15
Air Blank	0.000	21:15
Control Test	0.081	21:16
Air Blank	0.000	21:16
Diagnostics Check	OK	21:16

Data Download Date/Time: 05/01/2019 12:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: HAZLETON

DATE/TIME OF TEST: 04/21/2019 00:10:30

Diagnostics Check	OK	23:59	
Air Blank	0.000	23:59	
Control Test	0.080	23:59	
Air Blank	0.000	00:00	
Subject Sample	# 1	00:03	No Sample Provided
Breath Volume	0.000	00:03	
Air Blank	0.000	00:03	
Air Blank	0.000	00:05	
Subject Sample	# 2	00:09	No Sample Provided
Breath Volume	0.000	00:09	
Air Blank	0.000	00:09	
Control Test	0.080	00:09	
Air Blank	0.000	00:10	
Diagnostics Check	OK	00:10	

Data Download Date/Time: 05/01/2019 12:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: HAZLETON

DATE/TIME OF TEST: 04/21/2019 00:18:41

Diagnostics Check	OK	00:12
Air Blank	0.000	00:12
Control Test	0.080	00:13
Air Blank	0.000	00:13
Subject Sample # 1	0.136	00:14
Breath Volume	1.687	00:14
Air Blank	0.000	00:14
Air Blank	0.000	00:16
Subject Sample # 2	0.137	00:17
Breath Volume	2.210	00:17
Air Blank	0.000	00:17
Control Test	0.080	00:18
Air Blank	0.000	00:18
Diagnostics Check	OK	00:18

Data Download Date/Time: 05/01/2019 12:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: MAXWELL

DATE/TIME OF TEST: 04/21/2019 02:17:00

Diagnostics Check	OK	02:14	
Air Blank	0.000	02:14	
Control Test	0.080	02:14	
Air Blank	0.000	02:15	
Subject Sample	# 1	02:15	Subject Test Refused
Breath Volume	0.000	02:15	
Air Blank	0.000	02:16	
Control Test	0.080	02:16	
Air Blank	0.000	02:16	
Diagnostics Check	OK	02:17	

Data Download Date/Time: 05/01/2019 12:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:39

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 04/21/2019 21:06:19

Diagnostics Check	OK	21:00
Air Blank	0.000	21:00
Control Test	0.081	21:00
Air Blank	0.000	21:01
Subject Sample # 1	0.179	21:01
Breath Volume	2.851	21:01
Air Blank	0.000	21:02
Air Blank	0.000	21:04
Subject Sample # 2	0.174	21:04
Breath Volume	2.656	21:04
Air Blank	0.000	21:05
Control Test	0.079	21:05
Air Blank	0.000	21:06
Diagnostics Check	OK	21:06

Data Download Date/Time: 05/01/2019 12:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DIAZ,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:11

SUBJECT'S LAST NAME: CREMEANS

DATE/TIME OF TEST: 04/21/2019 22:38:34

Diagnostics Check	OK	22:31
Air Blank	0.000	22:32
Control Test	0.081	22:32
Air Blank	0.000	22:33
Subject Sample # 1	0.000	22:33
Breath Volume	3.042	22:33
Air Blank	0.000	22:34
Air Blank	0.000	22:36
Subject Sample # 2	0.000	22:37
Breath Volume	2.980	22:37
Air Blank	0.000	22:37
Control Test	0.080	22:37
Air Blank	0.000	22:38
Diagnostics Check	OK	22:38

Data Download Date/Time: 05/01/2019 12:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CAMACHO,K
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 04/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: CLANCE

DATE/TIME OF TEST: 04/22/2019 02:16:41

Diagnostics Check	OK	02:09
Air Blank	0.000	02:09
Control Test	0.081	02:10
Air Blank	0.000	02:10
Subject Sample # 1	0.225	02:11
Breath Volume	1.121	02:11
Air Blank	0.000	02:12
Air Blank	0.000	02:13
Subject Sample # 2	0.222	02:15
Breath Volume	1.355	02:15
Air Blank	0.000	02:15
Control Test	0.080	02:16
Air Blank	0.000	02:16
Diagnostics Check	OK	02:16

Data Download Date/Time: 05/01/2019 12:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 04/22/2019 23:20:49

Diagnostics Check	OK	23:14
Air Blank	0.000	23:14
Control Test	0.080	23:15
Air Blank	0.000	23:15
Subject Sample # 1	0.065	23:16
Breath Volume	1.203	23:16
Air Blank	0.000	23:16
Air Blank	0.000	23:18
Subject Sample # 2	0.064	23:19
Breath Volume	1.855	23:19
Air Blank	0.000	23:19
Control Test	0.080	23:20
Air Blank	0.000	23:20
Diagnostics Check	OK	23:20

Data Download Date/Time: 05/01/2019 12:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DIAZ,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 04/23/2019 04:59:49

Diagnostics Check	OK	04:53
Air Blank	0.000	04:53
Control Test	0.080	04:54
Air Blank	0.000	04:54
Subject Sample # 1	0.181	04:55
Breath Volume	2.472	04:55
Air Blank	0.000	04:55
Air Blank	0.000	04:57
Subject Sample # 2	0.183	04:58
Breath Volume	2.636	04:58
Air Blank	0.000	04:58
Control Test	0.079	04:59
Air Blank	0.000	04:59
Diagnostics Check	OK	04:59

Data Download Date/Time: 05/01/2019 12:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SOSA,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 04/24/2019 03:36:35

Diagnostics Check	OK	03:30
Air Blank	0.000	03:31
Control Test	0.080	03:31
Air Blank	0.000	03:31
Subject Sample # 1	0.000	03:32
Breath Volume	1.359	03:32
Air Blank	0.000	03:32
Air Blank	0.000	03:34
Subject Sample # 2	0.000	03:35
Breath Volume	2.621	03:35
Air Blank	0.000	03:35
Control Test	0.080	03:35
Air Blank	0.000	03:36
Diagnostics Check	OK	03:36

Data Download Date/Time: 05/01/2019 12:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS, CHRISTOPHER, D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON, C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: WITHERSPOON

DATE/TIME OF TEST: 04/26/2019 01:51:21

Diagnostics Check	OK	01:45
Air Blank	0.000	01:45
Control Test	0.081	01:45
Air Blank	0.000	01:46
Subject Sample # 1	0.074	01:46
Breath Volume	3.066	01:46
Air Blank	0.000	01:47
Air Blank	0.000	01:49
Subject Sample # 2	0.073	01:49
Breath Volume	3.683	01:49
Air Blank	0.000	01:50
Control Test	0.080	01:50
Air Blank	0.000	01:51
Diagnostics Check	OK	01:51

Data Download Date/Time: 05/01/2019 12:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BEALL,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: HILLS

DATE/TIME OF TEST: 04/27/2019 00:08:48

Diagnostics Check	OK	00:05	
Air Blank	0.000	00:06	
Control Test	0.079	00:06	
Air Blank	0.000	00:07	
Subject Sample	# 1	00:07	Subject Test Refused
Breath Volume	0.000	00:07	
Air Blank	0.000	00:07	
Control Test	0.079	00:08	
Air Blank	0.000	00:08	
Diagnostics Check	OK	00:08	

Data Download Date/Time: 05/01/2019 12:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS, CHRISTOPHER, D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JENSEN, D
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 04/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: DENTON

DATE/TIME OF TEST: 04/27/2019 01:41:12

Diagnostics Check	OK	01:23	
Air Blank	0.000	01:24	
Control Test	0.079	01:24	
Air Blank	0.000	01:25	
Subject Sample	# 1 0.203	01:28	Volume Not Met
Breath Volume	0.332	01:28	
Air Blank	0.000	01:29	
Air Blank	0.000	01:30	
Subject Sample	# 2 0.265	01:33	
Breath Volume	1.429	01:33	
Air Blank	0.000	01:34	
Air Blank	0.000	01:36	
Subject Sample	# 3 0.268	01:39	Volume Not Met
Breath Volume	0.441	01:39	
Air Blank	0.000	01:40	
Control Test	0.075	01:40	
Air Blank	0.000	01:41	
Diagnostics Check	OK	01:41	

Data Download Date/Time: 05/01/2019 12:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS, CHRISTOPHER, D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TORRES, W
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 04/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: COLGROVE

DATE/TIME OF TEST: 04/27/2019 02:51:20

Diagnostics Check	OK	02:45
Air Blank	0.000	02:45
Control Test	0.078	02:45
Air Blank	0.000	02:46
Subject Sample # 1	0.048	02:46
Breath Volume	3.441	02:46
Air Blank	0.000	02:47
Air Blank	0.000	02:49
Subject Sample # 2	0.046	02:49
Breath Volume	2.019	02:49
Air Blank	0.000	02:50
Control Test	0.079	02:50
Air Blank	0.000	02:51
Diagnostics Check	OK	02:51

Data Download Date/Time: 05/01/2019 12:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CAMACHO,F
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: KOLODINSKY

DATE/TIME OF TEST: 04/27/2019 03:43:34

Diagnostics Check	OK	03:40	
Air Blank	0.000	03:41	
Control Test	0.080	03:41	
Air Blank	0.000	03:41	
Subject Sample	# 1	03:42	Subject Test Refused
Breath Volume	0.000	03:42	
Air Blank	0.000	03:42	
Control Test	0.080	03:42	
Air Blank	0.000	03:43	
Diagnostics Check	OK	03:43	

Data Download Date/Time: 05/01/2019 12:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 04/27/2019 20:17:06

Diagnostics Check	OK	20:10
Air Blank	0.000	20:11
Control Test	0.080	20:11
Air Blank	0.000	20:11
Subject Sample # 1	0.191	20:12
Breath Volume	2.882	20:12
Air Blank	0.000	20:13
Air Blank	0.000	20:14
Subject Sample # 2	0.181	20:15
Breath Volume	2.871	20:15
Air Blank	0.000	20:16
Control Test	0.079	20:16
Air Blank	0.000	20:17
Diagnostics Check	OK	20:17

Data Download Date/Time: 05/01/2019 12:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CAMACHO,K
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 04/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:53

SUBJECT'S LAST NAME: MEUTEN

DATE/TIME OF TEST: 04/27/2019 21:19:53

Diagnostics Check	OK	21:13
Air Blank	0.000	21:13
Control Test	0.080	21:14
Air Blank	0.000	21:14
Subject Sample # 1	0.093	21:15
Breath Volume	3.710	21:15
Air Blank	0.000	21:15
Air Blank	0.000	21:17
Subject Sample # 2	0.093	21:18
Breath Volume	4.164	21:18
Air Blank	0.000	21:18
Control Test	0.080	21:19
Air Blank	0.000	21:19
Diagnostics Check	OK	21:19

Data Download Date/Time: 05/01/2019 12:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: ROSE

DATE/TIME OF TEST: 04/27/2019 23:47:11

Diagnostics Check	OK	23:44	
Air Blank	0.000	23:44	
Control Test	0.080	23:45	
Air Blank	0.000	23:45	
Subject Sample	# 1	23:45	Subject Test Refused
Breath Volume	0.000	23:45	
Air Blank	0.000	23:46	
Control Test	0.080	23:46	
Air Blank	0.000	23:47	
Diagnostics Check	OK	23:47	

Data Download Date/Time: 05/01/2019 12:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: INGERSOLL

DATE/TIME OF TEST: 04/28/2019 01:41:00

Diagnostics Check	OK	01:38	
Air Blank	0.000	01:38	
Control Test	0.080	01:38	
Air Blank	0.000	01:39	
Subject Sample	# 1	01:39	Subject Test Refused
Breath Volume	0.000	01:39	
Air Blank	0.000	01:40	
Control Test	0.080	01:40	
Air Blank	0.000	01:40	
Diagnostics Check	OK	01:41	

Data Download Date/Time: 05/01/2019 12:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DANZIE,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 04/28/2019 03:47:33

Diagnostics Check	OK	03:41
Air Blank	0.000	03:41
Control Test	0.080	03:41
Air Blank	0.000	03:42
Subject Sample # 1	0.156	03:43
Breath Volume	2.992	03:43
Air Blank	0.000	03:43
Air Blank	0.000	03:45
Subject Sample # 2	0.165	03:45
Breath Volume	2.656	03:45
Air Blank	0.000	03:46
Control Test	0.079	03:46
Air Blank	0.000	03:47
Diagnostics Check	OK	03:47

Data Download Date/Time: 05/01/2019 12:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: NOBORE,O
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 05:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:36

SUBJECT'S LAST NAME: VALVERDE

DATE/TIME OF TEST: 04/28/2019 05:57:41

Diagnostics Check	OK	05:57	
Air Blank		05:57	RFI Detect
Air Blank	0.000	05:57	

Data Download Date/Time: 05/01/2019 12:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: NOBORE,O
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 05:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:36

SUBJECT'S LAST NAME: GRANADA VALVERDE

DATE/TIME OF TEST: 04/28/2019 06:06:11

Diagnostics Check	OK	05:59
Air Blank	0.000	06:00
Control Test	0.080	06:00
Air Blank	0.000	06:01
Subject Sample # 1	0.143	06:01
Breath Volume	2.144	06:01
Air Blank	0.000	06:02
Air Blank	0.000	06:04
Subject Sample # 2	0.135	06:04
Breath Volume	3.316	06:04
Air Blank	0.000	06:05
Control Test	0.079	06:05
Air Blank	0.000	06:06
Diagnostics Check	OK	06:06

Data Download Date/Time: 05/01/2019 12:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HOGAN,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 04/29/2019 02:40:59

Diagnostics Check	OK	02:37	
Air Blank	0.000	02:37	
Control Test	0.080	02:37	
Air Blank	0.000	02:38	
Subject Sample	# 1	02:39	Subject Test Refused
Breath Volume	0.000	02:39	
Air Blank	0.000	02:40	
Control Test	0.080	02:40	
Air Blank	0.000	02:40	
Diagnostics Check	OK	02:40	

Data Download Date/Time: 05/01/2019 12:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CAMACHO,K
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 04/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: BOOTH

DATE/TIME OF TEST: 04/30/2019 02:17:21

Diagnostics Check	OK	02:09
Air Blank	0.000	02:10
Control Test	0.081	02:10
Air Blank	0.000	02:11
Subject Sample # 1	0.112	02:12
Breath Volume	2.550	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:14
Subject Sample # 2	0.118	02:15
Breath Volume	2.468	02:15
Air Blank	0.000	02:16
Control Test	0.080	02:16
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

Data Download Date/Time: 05/01/2019 12:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 04/30/2019 23:43:08

Diagnostics Check	OK	23:36
Air Blank	0.000	23:37
Control Test	0.079	23:37
Air Blank	0.000	23:38
Subject Sample # 1	0.210	23:38
Breath Volume	2.902	23:38
Air Blank	0.000	23:39
Air Blank	0.000	23:41
Subject Sample # 2	0.208	23:41
Breath Volume	2.796	23:41
Air Blank	0.000	23:42
Control Test	0.080	23:42
Air Blank	0.000	23:43
Diagnostics Check	OK	23:43

Data Download Date/Time: 05/01/2019 12:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASLINE,T'NEISHA,S
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: MILLS,D
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 03/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: CATALANO

DATE/TIME OF TEST: 04/04/2019 03:54:17

Diagnostics Check	OK	03:44
Air Blank	0.000	03:44
Control Test	0.080	03:45
Air Blank	0.000	03:45
Subject Sample # 1	0.170	03:46
Breath Volume	1.218	03:46
Air Blank	0.000	03:46
Air Blank	0.000	03:48
Subject Sample # 2	0.203	03:49
Breath Volume	1.460	03:49
Air Blank	0.000	03:50
Air Blank	0.000	03:51
Subject Sample # 3	0.188	03:52
Breath Volume	1.949	03:52
Air Blank	0.000	03:53
Control Test	0.080	03:53
Air Blank	0.000	03:54
Diagnostics Check	OK	03:54

Data Download Date/Time: 04/21/2019 07:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARCHER,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: BAUMGARTNER,NATHANIAL
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 10:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 04/04/2019 15:36:20

Diagnostics Check	OK	15:30
Air Blank	0.000	15:30
Control Test	0.080	15:31
Air Blank	0.000	15:31
Subject Sample # 1	0.000	15:32
Breath Volume	1.421	15:32
Air Blank	0.000	15:32
Air Blank	0.000	15:34
Subject Sample # 2	0.000	15:34
Breath Volume	1.125	15:34
Air Blank	0.000	15:35
Control Test	0.080	15:35
Air Blank	0.000	15:36
Diagnostics Check	OK	15:36

Data Download Date/Time: 04/21/2019 07:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNOW,C,D
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: SNOW,C
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: ADMIN
ARREST TIME: 09:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: MANNING

DATE/TIME OF TEST: 04/06/2019 09:30:25

Diagnostics Check	OK	09:29	
Air Blank	0.000	09:29	
Control Test	0.080	09:29	
Air Blank		09:29	RFI Detect
Air Blank	0.000	09:30	

Data Download Date/Time: 04/21/2019 07:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNOW,C,D
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: SNOW,C
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: ADMIN
ARREST TIME: 09:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: MANNING

DATE/TIME OF TEST: 04/06/2019 09:39:00

Diagnostics Check	OK	09:32
Air Blank	0.000	09:32
Control Test	0.080	09:32
Air Blank	0.000	09:33
Subject Sample # 1	0.266	09:34
Breath Volume	1.402	09:34
Air Blank	0.000	09:34
Air Blank	0.000	09:36
Subject Sample # 2	0.248	09:37
Breath Volume	1.351	09:37
Air Blank	0.000	09:38
Control Test	0.078	09:38
Air Blank	0.000	09:38
Diagnostics Check	OK	09:39

Data Download Date/Time: 04/21/2019 07:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNOW,C,D
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: SNOW,C
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: ADMIN
ARREST TIME: 09:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:15

SUBJECT'S LAST NAME: DUNLAP

DATE/TIME OF TEST: 04/06/2019 09:48:55

Diagnostics Check	OK	09:42
Air Blank	0.000	09:43
Control Test	0.079	09:43
Air Blank	0.000	09:44
Subject Sample # 1	0.000	09:44
Breath Volume	2.671	09:44
Air Blank	0.000	09:45
Air Blank	0.000	09:47
Subject Sample # 2	0.000	09:47
Breath Volume	2.765	09:47
Air Blank	0.000	09:48
Control Test	0.079	09:48
Air Blank	0.000	09:48
Diagnostics Check	OK	09:48

Data Download Date/Time: 04/21/2019 07:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENTZ,JARROD,C
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WHITE,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: GARDNER

DATE/TIME OF TEST: 04/16/2019 00:17:51

Diagnostics Check	OK	00:11
Air Blank	0.000	00:11
Control Test	0.080	00:12
Air Blank	0.000	00:12
Subject Sample # 1	0.000	00:13
Breath Volume	2.171	00:13
Air Blank	0.000	00:13
Air Blank	0.000	00:15
Subject Sample # 2	0.000	00:16
Breath Volume	2.769	00:16
Air Blank	0.000	00:16
Control Test	0.078	00:17
Air Blank	0.000	00:17
Diagnostics Check	OK	00:17

Data Download Date/Time: 04/21/2019 07:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENTZ,JARROD,C
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: ADAMS,S
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: SANFORD

DATE/TIME OF TEST: 04/20/2019 23:20:51

Diagnostics Check	OK	23:12
Air Blank	0.000	23:13
Control Test	0.080	23:13
Air Blank	0.000	23:14
Subject Sample # 1	0.135	23:15
Breath Volume	1.375	23:15
Air Blank	0.000	23:16
Air Blank	0.000	23:17
Subject Sample # 2	0.132	23:19
Breath Volume	1.445	23:19
Air Blank	0.000	23:19
Control Test	0.079	23:20
Air Blank	0.000	23:20
Diagnostics Check	OK	23:20

Data Download Date/Time: 04/21/2019 07:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUNNER,JOSEPH,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: RUNNER,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: REESE

DATE/TIME OF TEST: 04/02/2019 00:16:21

Diagnostics Check	OK	00:13	
Air Blank	0.000	00:13	
Control Test	0.079	00:13	
Air Blank	0.000	00:14	
Subject Sample	# 1	00:14	RFI Detect
Breath Volume	0.000	00:14	
Air Blank	0.000	00:15	
Control Test	0.080	00:15	
Air Blank	0.000	00:16	
Diagnostics Check	OK	00:16	

Data Download Date/Time: 04/27/2019 11:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUNNER,JOSEPH,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: RUNNER,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: REESE

DATE/TIME OF TEST: 04/02/2019 00:30:15

Diagnostics Check	OK	00:24
Air Blank	0.000	00:24
Control Test	0.080	00:24
Air Blank	0.000	00:25
Subject Sample # 1	0.276	00:25
Breath Volume	1.609	00:25
Air Blank	0.000	00:26
Air Blank	0.000	00:28
Subject Sample # 2	0.263	00:28
Breath Volume	1.527	00:28
Air Blank	0.000	00:29
Control Test	0.079	00:29
Air Blank	0.000	00:30
Diagnostics Check	OK	00:30

Data Download Date/Time: 04/27/2019 11:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUCHANAN,SUMMER,N
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BUCHANAN,SUMMER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: GUZMAN

DATE/TIME OF TEST: 04/20/2019 03:26:56

Diagnostics Check	OK	03:18	
Air Blank	0.000	03:18	
Control Test	0.079	03:18	
Air Blank	0.000	03:19	
Subject Sample	# 1	03:22	No Sample Provided
Breath Volume	0.000	03:22	
Air Blank	0.000	03:23	
Air Blank	0.000	03:25	
Subject Sample	# 2	03:25	Subject Test Refused
Breath Volume	0.000	03:25	
Air Blank	0.000	03:26	
Control Test	0.080	03:26	
Air Blank	0.000	03:26	
Diagnostics Check	OK	03:26	

Data Download Date/Time: 04/27/2019 11:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTIZ,RICARDO,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: MARTINEZ,EDGAR
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: LILLY

DATE/TIME OF TEST: 04/20/2019 21:38:44

Diagnostics Check	OK	21:30	
Air Blank	0.000	21:30	
Control Test	0.082	21:31	
Air Blank	0.000	21:31	
Subject Sample	# 1	21:34	No Sample Provided
Breath Volume	0.000	21:34	
Air Blank	0.000	21:35	
Air Blank	0.000	21:37	
Subject Sample	# 2	21:37	Subject Test Refused
Breath Volume	0.000	21:37	
Air Blank	0.000	21:37	
Control Test	0.081	21:38	
Air Blank	0.000	21:38	
Diagnostics Check	OK	21:38	

Data Download Date/Time: 04/27/2019 11:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONNELL,JACOB,B
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: CONNELL,JACOB
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: MUNLEY

DATE/TIME OF TEST: 04/23/2019 03:27:50

Diagnostics Check	OK	03:21
Air Blank	0.000	03:21
Control Test	0.078	03:22
Air Blank	0.000	03:22
Subject Sample # 1	0.260	03:23
Breath Volume	2.679	03:23
Air Blank	0.000	03:24
Air Blank	0.000	03:25
Subject Sample # 2	0.254	03:26
Breath Volume	2.464	03:26
Air Blank	0.000	03:26
Control Test	0.077	03:27
Air Blank	0.000	03:27
Diagnostics Check	OK	03:27

Data Download Date/Time: 04/27/2019 11:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHAUT,RICHARD,J
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: SCHAUT,RICHARD
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: DERRICK

DATE/TIME OF TEST: 04/17/2019 22:23:52

Diagnostics Check	OK	22:13
Air Blank	0.000	22:13
Control Test	0.079	22:14
Air Blank	0.000	22:14
Subject Sample # 1	0.206	22:15
Breath Volume	1.453	22:15
Air Blank	0.000	22:16
Air Blank	0.000	22:18
Subject Sample # 2	0.231	22:19
Breath Volume	1.640	22:19
Air Blank	0.000	22:19
Air Blank	0.000	22:21
Subject Sample # 3	0.210	22:22
Breath Volume	1.722	22:22
Air Blank	0.000	22:22
Control Test	0.076	22:23
Air Blank	0.000	22:23
Diagnostics Check	OK	22:23

Data Download Date/Time: 04/25/2019 10:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAYE, MAKINA,
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: CAJUSTE, A
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 04/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 04/30/2019 00:03:58

Diagnostics Check	OK	23:50	
Air Blank	0.000	23:50	
Control Test	0.080	23:51	
Air Blank	0.000	23:51	
Subject Sample	# 2	23:54	No Sample Provided
Breath Volume	0.000	23:54	
Air Blank	0.000	23:55	
Air Blank	0.000	23:57	
Subject Sample	# 3	0.113	23:58
Breath Volume	2.527	23:58	
Air Blank	0.000	23:59	
Air Blank	0.000	00:01	
Subject Sample	# 1	0.109	00:02
Breath Volume	1.871	00:02	
Air Blank	0.000	00:03	
Control Test	0.078	00:03	
Air Blank	0.000	00:03	
Diagnostics Check	OK	00:03	

Data Download Date/Time: 05/30/2019 15:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRONZ,BRIAN,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,KRISTEN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:42

SUBJECT'S LAST NAME: HOPKINS

DATE/TIME OF TEST: 04/02/2019 21:06:13

Diagnostics Check	OK	21:02	
Air Blank	0.000	21:03	
Control Test	0.079	21:03	
Air Blank	0.000	21:04	
Subject Sample	# 1	21:04	Interferent Detect
Breath Volume	2.492	21:04	
Air Blank	0.000	21:05	
Control Test	0.079	21:05	
Air Blank	0.000	21:06	
Diagnostics Check	OK	21:06	

Data Download Date/Time: 04/10/2019 08:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRONZ,BRIAN,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,BRIAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:42

SUBJECT'S LAST NAME: HOPKINS

DATE/TIME OF TEST: 04/02/2019 21:26:20

Diagnostics Check	OK	21:22	
Air Blank	0.000	21:22	
Control Test	0.078	21:23	
Air Blank	0.000	21:23	
Subject Sample	# 1	21:24	Interferent Detect
Breath Volume	3.031	21:24	
Air Blank	0.000	21:25	
Control Test	0.079	21:25	
Air Blank	0.000	21:26	
Diagnostics Check	OK	21:26	

Data Download Date/Time: 04/10/2019 08:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUM,ZACKARY,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KARAYIANES,M
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: MUREB

DATE/TIME OF TEST: 04/04/2019 23:19:03

Diagnostics Check	OK	23:13
Air Blank	0.000	23:13
Control Test	0.080	23:13
Air Blank	0.000	23:14
Subject Sample # 1	0.000	23:14
Breath Volume	3.679	23:14
Air Blank	0.000	23:15
Air Blank	0.000	23:17
Subject Sample # 2	0.000	23:17
Breath Volume	4.128	23:17
Air Blank	0.000	23:18
Control Test	0.079	23:18
Air Blank	0.000	23:18
Diagnostics Check	OK	23:19

Data Download Date/Time: 04/10/2019 08:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LAMOUR,ZACHARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 04/05/2019 00:15:53

Diagnostics Check	OK	00:09
Air Blank	0.000	00:10
Control Test	0.078	00:10
Air Blank	0.000	00:10
Subject Sample # 1	0.266	00:11
Breath Volume	3.109	00:11
Air Blank	0.000	00:12
Air Blank	0.000	00:13
Subject Sample # 2	0.264	00:14
Breath Volume	2.960	00:14
Air Blank	0.000	00:14
Control Test	0.078	00:15
Air Blank	0.000	00:15
Diagnostics Check	OK	00:15

Data Download Date/Time: 04/10/2019 08:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE, RONALD,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SMITH, EMILY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:26

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 04/05/2019 16:55:59

Diagnostics Check	OK	16:47	
Air Blank	0.000	16:47	
Control Test	0.078	16:48	
Air Blank	0.000	16:48	
Subject Sample # 1	0.257	16:49	
Breath Volume	1.582	16:49	
Air Blank	0.000	16:50	
Air Blank	0.000	16:51	
Subject Sample # 2		16:54	Subject Test Refused
Breath Volume	0.074	16:54	
Air Blank	0.000	16:55	
Control Test	0.078	16:55	
Air Blank	0.000	16:55	
Diagnostics Check	OK	16:55	

Data Download Date/Time: 04/10/2019 08:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LAMOUR,ZACHARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: MALDONADO-AVILA

DATE/TIME OF TEST: 04/06/2019 00:01:31

Diagnostics Check	OK	23:55
Air Blank	0.000	23:56
Control Test	0.078	23:56
Air Blank	0.000	23:56
Subject Sample # 1	0.188	23:57
Breath Volume	1.832	23:57
Air Blank	0.000	23:57
Air Blank	0.000	23:59
Subject Sample # 2	0.208	23:59
Breath Volume	3.363	23:59
Air Blank	0.000	00:00
Control Test	0.077	00:00
Air Blank	0.000	00:01
Diagnostics Check	OK	00:01

Data Download Date/Time: 04/10/2019 08:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINES,PAUL,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HINES,PAUL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:47

SUBJECT'S LAST NAME: LORD

DATE/TIME OF TEST: 04/06/2019 17:25:42

Diagnostics Check	OK	17:11	
Air Blank	0.000	17:11	
Control Test	0.077	17:12	
Air Blank	0.000	17:12	
Subject Sample	# 1	17:15	No Sample Provided
Breath Volume	0.000	17:15	
Air Blank	0.000	17:16	
Air Blank	0.000	17:18	
Subject Sample	# 2	0.065	17:20
Breath Volume	1.101	17:20	
Air Blank	0.000	17:21	
Air Blank	0.000	17:23	
Subject Sample	# 3	0.066	17:24
Breath Volume	1.816	17:24	
Air Blank	0.000	17:24	
Control Test	0.078	17:25	
Air Blank	0.000	17:25	
Diagnostics Check	OK	17:25	

Data Download Date/Time: 04/10/2019 08:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,BRENT,D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CLARK,BRENT
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: DASHNER NATHAN

DATE/TIME OF TEST: 04/06/2019 20:52:36

Diagnostics Check	OK	20:46
Air Blank	0.000	20:46
Control Test	0.078	20:46
Air Blank	0.000	20:47
Subject Sample # 1	0.000	20:47
Breath Volume	2.843	20:47
Air Blank	0.000	20:48
Air Blank	0.000	20:50
Subject Sample # 2	0.000	20:51
Breath Volume	1.750	20:51
Air Blank	0.000	20:51
Control Test	0.079	20:52
Air Blank	0.000	20:52
Diagnostics Check	OK	20:52

Data Download Date/Time: 04/10/2019 08:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAIA,MICHAEL,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HINES,PAUL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: NEWKIRK

DATE/TIME OF TEST: 04/07/2019 03:19:24

Diagnostics Check	OK	03:13
Air Blank	0.000	03:13
Control Test	0.077	03:14
Air Blank	0.000	03:14
Subject Sample # 1	0.103	03:15
Breath Volume	1.652	03:15
Air Blank	0.000	03:15
Air Blank	0.000	03:17
Subject Sample # 2	0.101	03:17
Breath Volume	2.089	03:17
Air Blank	0.000	03:18
Control Test	0.079	03:18
Air Blank	0.000	03:19
Diagnostics Check	OK	03:19

Data Download Date/Time: 04/10/2019 08:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,BRENT,D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CLARK,BRENT
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: VANDENBURGH

DATE/TIME OF TEST: 04/07/2019 22:21:40

Diagnostics Check	OK	22:14	
Air Blank	0.000	22:15	
Control Test	0.076	22:15	
Air Blank	0.000	22:16	
Subject Sample # 1	0.152	22:16	
Breath Volume	3.265	22:16	
Air Blank	0.000	22:17	
Air Blank	0.000	22:19	
Subject Sample # 2	0.141	22:20	
Breath Volume	2.089	22:20	
Air Blank	0.000	22:20	
Control Test	0.070	22:21	Control Outside Tolerance
Air Blank	0.000	22:21	

Data Download Date/Time: 04/10/2019 08:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,BRENT,D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CLARK,BRENT
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: V535928470230

DATE/TIME OF TEST: 04/07/2019 22:31:29

Diagnostics Check	OK	22:24
Air Blank	0.000	22:25
Control Test	0.076	22:25
Air Blank	0.000	22:26
Subject Sample # 1	0.156	22:26
Breath Volume	2.808	22:26
Air Blank	0.000	22:27
Air Blank	0.000	22:29
Subject Sample # 2	0.151	22:30
Breath Volume	2.386	22:30
Air Blank	0.000	22:30
Control Test	0.076	22:30
Air Blank	0.000	22:31
Diagnostics Check	OK	22:31

Data Download Date/Time: 04/10/2019 08:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACRI,MICHEAL,N
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCCULLOUGH,SEAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 04/08/2019 01:26:14

Diagnostics Check	OK	01:16
Air Blank	0.000	01:17
Control Test	0.077	01:17
Air Blank	0.000	01:17
Subject Sample # 1	0.210	01:18
Breath Volume	1.671	01:18
Air Blank	0.000	01:19
Air Blank	0.000	01:21
Subject Sample # 2	0.252	01:21
Breath Volume	4.824	01:21
Air Blank	0.000	01:22
Air Blank	0.000	01:24
Subject Sample # 3	0.247	01:24
Breath Volume	4.085	01:24
Air Blank	0.000	01:25
Control Test	0.082	01:25
Air Blank	0.000	01:26
Diagnostics Check	OK	01:26

Data Download Date/Time: 04/10/2019 08:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINES,PAUL,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCCULLOUGH,SEAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: PHAKEOVILAY

DATE/TIME OF TEST: 04/09/2019 00:22:05

Diagnostics Check	OK	00:15
Air Blank	0.000	00:16
Control Test	0.079	00:16
Air Blank	0.000	00:17
Subject Sample # 1	0.196	00:17
Breath Volume	2.222	00:17
Air Blank	0.000	00:18
Air Blank	0.000	00:20
Subject Sample # 2	0.196	00:20
Breath Volume	2.285	00:20
Air Blank	0.000	00:21
Control Test	0.078	00:21
Air Blank	0.000	00:22
Diagnostics Check	OK	00:22

Data Download Date/Time: 04/10/2019 08:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISTI,JORDAN,A
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SISTI,JORDAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: LATIMORE

DATE/TIME OF TEST: 04/10/2019 00:48:21

Diagnostics Check	OK	00:41
Air Blank	0.000	00:42
Control Test	0.078	00:42
Air Blank	0.000	00:43
Subject Sample # 1	0.305	00:43
Breath Volume	2.031	00:43
Air Blank	0.000	00:44
Air Blank	0.000	00:46
Subject Sample # 2	0.313	00:46
Breath Volume	2.566	00:46
Air Blank	0.000	00:47
Control Test	0.077	00:47
Air Blank	0.000	00:48
Diagnostics Check	OK	00:48

Data Download Date/Time: 04/10/2019 08:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE, RONALD, W
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: L'ESPERANCE, DANIEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:54

SUBJECT'S LAST NAME: PAULK

DATE/TIME OF TEST: 04/12/2019 05:21:46

Diagnostics Check	OK	05:15
Air Blank	0.000	05:15
Control Test	0.079	05:15
Air Blank	0.000	05:16
Subject Sample # 1	0.193	05:17
Breath Volume	2.199	05:17
Air Blank	0.000	05:17
Air Blank	0.000	05:19
Subject Sample # 2	0.186	05:20
Breath Volume	1.855	05:20
Air Blank	0.000	05:20
Control Test	0.078	05:21
Air Blank	0.000	05:21
Diagnostics Check	OK	05:21

Data Download Date/Time: 05/09/2019 09:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,BRENT,D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CLARK,BRENT
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:21

SUBJECT'S LAST NAME: ALLMAN

DATE/TIME OF TEST: 04/14/2019 20:46:15

Diagnostics Check	OK	20:41	
Air Blank	0.000	20:42	
Control Test	0.079	20:42	
Air Blank	0.000	20:42	
Subject Sample	# 1	20:44	Subject Test Refused
Breath Volume	0.000	20:44	
Air Blank	0.000	20:45	
Control Test	0.077	20:45	
Air Blank	0.000	20:46	
Diagnostics Check	OK	20:46	

Data Download Date/Time: 05/09/2019 09:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACRI,MICHEAL,N
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SIREN,ERIC
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: MADRY

DATE/TIME OF TEST: 04/15/2019 00:47:16

Diagnostics Check	OK	00:35	
Air Blank	0.000	00:36	
Control Test	0.079	00:36	
Air Blank	0.000	00:37	
Subject Sample	# 1	00:40	No Sample Provided
Breath Volume	0.000	00:40	
Air Blank	0.000	00:40	
Air Blank	0.000	00:42	
Subject Sample	# 2	00:45	No Sample Provided
Breath Volume	0.000	00:45	
Air Blank	0.000	00:46	
Control Test	0.077	00:46	
Air Blank	0.000	00:47	
Diagnostics Check	OK	00:47	

Data Download Date/Time: 05/09/2019 09:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACRI, MICHEAL, N
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SIRENI, ERIC
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: MADRY

DATE/TIME OF TEST: 04/15/2019 00:52:09

Diagnostics Check	OK	00:49	
Air Blank	0.000	00:49	
Control Test	0.079	00:50	
Air Blank	0.000	00:50	
Subject Sample	# 1	00:50	Subject Test Refused
Breath Volume	0.000	00:50	
Air Blank	0.000	00:51	
Control Test	0.078	00:51	
Air Blank	0.000	00:52	
Diagnostics Check	OK	00:52	

Data Download Date/Time: 05/09/2019 09:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LESPERANCE,DANIEL,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LESPERANCE,DANIEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 04/20/2019 00:32:50

Diagnostics Check	OK	00:26
Air Blank	0.000	00:27
Control Test	0.079	00:27
Air Blank	0.000	00:27
Subject Sample # 1	0.180	00:28
Breath Volume	1.792	00:28
Air Blank	0.000	00:28
Air Blank	0.000	00:30
Subject Sample # 2	0.188	00:31
Breath Volume	2.535	00:31
Air Blank	0.000	00:31
Control Test	0.078	00:32
Air Blank	0.000	00:32
Diagnostics Check	OK	00:32

Data Download Date/Time: 05/09/2019 09:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LESPERANCE,DANIEL,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LAMOUR,ZACHARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: FACEY

DATE/TIME OF TEST: 04/21/2019 01:52:59

Diagnostics Check	OK	01:46
Air Blank	0.000	01:47
Control Test	0.079	01:47
Air Blank	0.000	01:47
Subject Sample # 1	0.203	01:48
Breath Volume	2.062	01:48
Air Blank	0.000	01:49
Air Blank	0.000	01:51
Subject Sample # 2	0.203	01:51
Breath Volume	3.003	01:51
Air Blank	0.000	01:52
Control Test	0.077	01:52
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

Data Download Date/Time: 05/09/2019 09:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LESPERANCE,DANIEL,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HINES,PAUL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: GONZALEZ ROBLES

DATE/TIME OF TEST: 04/21/2019 02:42:36

Diagnostics Check	OK	02:35
Air Blank	0.000	02:36
Control Test	0.079	02:36
Air Blank	0.000	02:37
Subject Sample # 1	0.000	02:38
Breath Volume	2.199	02:38
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample # 2	0.000	02:41
Breath Volume	2.058	02:41
Air Blank	0.000	02:41
Control Test	0.079	02:42
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 05/09/2019 09:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUM,ZACKARY,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LESPERANCE,D
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: MAGOS

DATE/TIME OF TEST: 04/24/2019 02:04:23

Diagnostics Check	OK	01:58
Air Blank	0.000	01:58
Control Test	0.078	01:58
Air Blank	0.000	01:59
Subject Sample # 1	0.253	01:59
Breath Volume	3.015	01:59
Air Blank	0.000	02:00
Air Blank	0.000	02:02
Subject Sample # 2	0.253	02:02
Breath Volume	2.867	02:02
Air Blank	0.000	02:03
Control Test	0.077	02:03
Air Blank	0.000	02:04
Diagnostics Check	OK	02:04

Data Download Date/Time: 05/09/2019 09:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WOOLVERTON,JUSTIN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: MCDOWELL

DATE/TIME OF TEST: 04/25/2019 04:11:02

Diagnostics Check	OK	04:02	
Air Blank	0.000	04:03	
Control Test	0.080	04:03	
Air Blank	0.000	04:03	
Subject Sample	# 1	04:06	No Sample Provided
Breath Volume	0.000	04:06	
Air Blank	0.000	04:07	
Air Blank	0.000	04:09	
Subject Sample	# 2	04:09	Subject Test Refused
Breath Volume	0.000	04:09	
Air Blank	0.000	04:10	
Control Test	0.079	04:10	
Air Blank	0.000	04:10	
Diagnostics Check	OK	04:11	

Data Download Date/Time: 05/09/2019 09:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAIA,MICHAEL,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCCULLOUGH,SEAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:04

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 04/25/2019 05:30:35

Diagnostics Check	OK	05:25	
Air Blank	0.000	05:25	
Control Test	0.078	05:25	
Air Blank	0.000	05:26	
Subject Sample	# 1	05:29	Subject Test Refused
Breath Volume	0.000	05:29	
Air Blank	0.000	05:29	
Control Test	0.080	05:30	
Air Blank	0.000	05:30	
Diagnostics Check	OK	05:30	

Data Download Date/Time: 05/09/2019 09:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LESPERANCE,DANIEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 04/27/2019 00:34:44

Diagnostics Check	OK	00:26
Air Blank	0.000	00:26
Control Test	0.080	00:27
Air Blank	0.000	00:27
Subject Sample # 1	0.054	00:30
Breath Volume	1.878	00:30
Air Blank	0.000	00:30
Air Blank	0.000	00:32
Subject Sample # 2	0.052	00:33
Breath Volume	1.738	00:33
Air Blank	0.000	00:33
Control Test	0.079	00:34
Air Blank	0.000	00:34
Diagnostics Check	OK	00:34

Data Download Date/Time: 05/09/2019 09:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,BRENT,D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CLARK,BRENT
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:27

SUBJECT'S LAST NAME: GREER

DATE/TIME OF TEST: 04/28/2019 15:58:40

Diagnostics Check	OK	15:48	
Air Blank	0.000	15:48	
Control Test	0.077	15:48	
Air Blank	0.000	15:49	
Subject Sample	# 1 0.419	15:50	No .020 agreement
Breath Volume	1.667	15:50	
Air Blank	0.000	15:51	
Air Blank	0.000	15:52	
Subject Sample	# 2 0.393	15:54	No .020 agreement
Breath Volume	1.449	15:54	
Air Blank	0.000	15:54	
Air Blank	0.000	15:56	
Subject Sample	# 3 0.349	15:57	No .020 agreement
Breath Volume	1.152	15:57	
Air Blank	0.000	15:57	
Control Test	0.075	15:58	
Air Blank	0.000	15:58	
Diagnostics Check	OK	15:58	

Data Download Date/Time: 05/09/2019 09:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,BRENT,D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CAPRA,JOHN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: BISHOP

DATE/TIME OF TEST: 04/28/2019 21:02:01

Diagnostics Check	OK	20:55
Air Blank	0.000	20:56
Control Test	0.078	20:56
Air Blank	0.000	20:57
Subject Sample # 1	0.231	20:57
Breath Volume	3.195	20:57
Air Blank	0.000	20:58
Air Blank	0.000	21:00
Subject Sample # 2	0.235	21:00
Breath Volume	3.144	21:00
Air Blank	0.000	21:01
Control Test	0.078	21:01
Air Blank	0.000	21:01
Diagnostics Check	OK	21:02

Data Download Date/Time: 05/09/2019 09:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,BRENT,D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HANEWINKEL,JAY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:06

SUBJECT'S LAST NAME: SEBASTIAN

DATE/TIME OF TEST: 04/28/2019 22:34:07

Diagnostics Check	OK	22:27
Air Blank	0.000	22:27
Control Test	0.077	22:27
Air Blank	0.000	22:28
Subject Sample # 1	0.171	22:29
Breath Volume	1.292	22:29
Air Blank	0.000	22:29
Air Blank	0.000	22:31
Subject Sample # 2	0.180	22:32
Breath Volume	2.171	22:32
Air Blank	0.000	22:33
Control Test	0.078	22:33
Air Blank	0.000	22:34
Diagnostics Check	OK	22:34

Data Download Date/Time: 05/09/2019 09:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAIA,MICHAEL,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WOOLVERTON,JUSTIN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 04/29/2019 01:04:23

Diagnostics Check	OK	00:57
Air Blank	0.000	00:58
Control Test	0.078	00:58
Air Blank	0.000	00:59
Subject Sample # 1	0.211	00:59
Breath Volume	2.410	00:59
Air Blank	0.000	01:00
Air Blank	0.000	01:02
Subject Sample # 2	0.203	01:02
Breath Volume	2.523	01:02
Air Blank	0.000	01:03
Control Test	0.077	01:03
Air Blank	0.000	01:04
Diagnostics Check	OK	01:04

Data Download Date/Time: 05/09/2019 09:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WEISKOPF,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 14:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:11

SUBJECT'S LAST NAME: KAY

DATE/TIME OF TEST: 04/30/2019 15:37:51

Diagnostics Check	OK	15:31
Air Blank	0.000	15:31
Control Test	0.079	15:32
Air Blank	0.000	15:32
Subject Sample # 1	0.296	15:33
Breath Volume	1.597	15:33
Air Blank	0.000	15:34
Air Blank	0.000	15:35
Subject Sample # 2	0.289	15:36
Breath Volume	1.566	15:36
Air Blank	0.000	15:36
Control Test	0.078	15:37
Air Blank	0.000	15:37
Diagnostics Check	OK	15:37

Data Download Date/Time: 05/09/2019 09:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,D,R
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: SANDIFER,R
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 04/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:23

SUBJECT'S LAST NAME: NORRIS

DATE/TIME OF TEST: 04/02/2019 18:52:30

Diagnostics Check	OK	18:45
Air Blank	0.000	18:45
Control Test	0.080	18:46
Air Blank	0.000	18:46
Subject Sample # 1	0.201	18:47
Breath Volume	1.617	18:47
Air Blank	0.000	18:48
Air Blank	0.000	18:50
Subject Sample # 2	0.189	18:51
Breath Volume	2.230	18:51
Air Blank	0.000	18:51
Control Test	0.079	18:51
Air Blank	0.000	18:52
Diagnostics Check	OK	18:52

Data Download Date/Time: 04/22/2019 09:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCracken,J,D
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: DELACRUZ,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 04/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 04/03/2019 22:51:19

Diagnostics Check	OK	22:48	
Air Blank	0.000	22:48	
Control Test	0.081	22:49	
Air Blank	0.000	22:49	
Subject Sample	# 1	22:49	Subject Test Refused
Breath Volume	0.000	22:49	
Air Blank	0.000	22:50	
Control Test	0.079	22:50	
Air Blank	0.000	22:51	
Diagnostics Check	OK	22:51	

Data Download Date/Time: 04/22/2019 09:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRES,J,L
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: SHOOK,J
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 04/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: MCCORMICK

DATE/TIME OF TEST: 04/09/2019 22:27:19

Diagnostics Check	OK	22:20
Air Blank	0.000	22:20
Control Test	0.080	22:21
Air Blank	0.000	22:21
Subject Sample # 1	0.174	22:22
Breath Volume	1.332	22:22
Air Blank	0.000	22:23
Air Blank	0.000	22:24
Subject Sample # 2	0.179	22:25
Breath Volume	1.898	22:25
Air Blank	0.000	22:26
Control Test	0.079	22:26
Air Blank	0.000	22:27
Diagnostics Check	OK	22:27

Data Download Date/Time: 04/22/2019 09:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KUTZ,T,E
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 04/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: SVOBODA

DATE/TIME OF TEST: 04/14/2019 03:19:06

Diagnostics Check	OK	03:16	
Air Blank	0.000	03:16	
Control Test	0.081	03:17	
Air Blank	0.000	03:17	
Subject Sample	# 1	03:17	Subject Test Refused
Breath Volume	0.000	03:17	
Air Blank	0.000	03:18	
Control Test	0.079	03:18	
Air Blank	0.000	03:19	
Diagnostics Check	OK	03:19	

Data Download Date/Time: 04/22/2019 09:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,E,E
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: SHOOK,J
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 04/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: TILL

DATE/TIME OF TEST: 04/14/2019 04:05:51

Diagnostics Check	OK	04:03	
Air Blank	0.000	04:03	
Control Test	0.079	04:03	
Air Blank	0.000	04:04	
Subject Sample	# 1	04:04	Subject Test Refused
Breath Volume	0.000	04:04	
Air Blank	0.000	04:04	
Control Test	0.080	04:05	
Air Blank	0.000	04:05	
Diagnostics Check	OK	04:05	

Data Download Date/Time: 04/22/2019 09:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,E,E
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: DELACRUZ,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 04/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: CANNATA

DATE/TIME OF TEST: 04/14/2019 22:48:48

Diagnostics Check	OK	22:39
Air Blank	0.000	22:40
Control Test	0.080	22:40
Air Blank	0.000	22:40
Subject Sample # 1	0.269	22:43
Breath Volume	2.160	22:43
Air Blank	0.000	22:44
Air Blank	0.000	22:46
Subject Sample # 2	0.253	22:47
Breath Volume	1.523	22:47
Air Blank	0.000	22:47
Control Test	0.079	22:48
Air Blank	0.000	22:48
Diagnostics Check	OK	22:48

Data Download Date/Time: 04/22/2019 09:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KUTZ,T,E
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: SHOOK,J
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 04/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: BEASLEY

DATE/TIME OF TEST: 04/15/2019 02:57:50

Diagnostics Check	OK	02:51
Air Blank	0.000	02:52
Control Test	0.080	02:52
Air Blank	0.000	02:52
Subject Sample # 1	0.138	02:53
Breath Volume	2.183	02:53
Air Blank	0.000	02:54
Air Blank	0.000	02:55
Subject Sample # 2	0.142	02:56
Breath Volume	2.550	02:56
Air Blank	0.000	02:56
Control Test	0.079	02:57
Air Blank	0.000	02:57
Diagnostics Check	OK	02:57

Data Download Date/Time: 04/22/2019 09:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCracken,J,D
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 04/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: TUCKER

DATE/TIME OF TEST: 04/15/2019 03:23:14

Diagnostics Check	OK	03:20	
Air Blank	0.000	03:20	
Control Test	0.081	03:21	
Air Blank	0.000	03:21	
Subject Sample	# 1	03:21	Subject Test Refused
Breath Volume	0.000	03:21	
Air Blank	0.000	03:22	
Control Test	0.080	03:22	
Air Blank	0.000	03:23	
Diagnostics Check	OK	03:23	

Data Download Date/Time: 04/22/2019 09:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCracken,J,D
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: DILL,T
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 04/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 04/15/2019 04:02:08

Diagnostics Check	OK	03:55
Air Blank	0.000	03:56
Control Test	0.081	03:56
Air Blank	0.000	03:57
Subject Sample # 1	0.180	03:57
Breath Volume	1.875	03:57
Air Blank	0.000	03:58
Air Blank	0.000	04:00
Subject Sample # 2	0.190	04:00
Breath Volume	1.156	04:00
Air Blank	0.000	04:01
Control Test	0.080	04:01
Air Blank	0.000	04:02
Diagnostics Check	OK	04:02

Data Download Date/Time: 04/22/2019 09:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOOLARD,R,C
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: PADGETT,W
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 04/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 11:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:36

SUBJECT'S LAST NAME: PADGETT

DATE/TIME OF TEST: 04/15/2019 13:06:31

Diagnostics Check	OK	13:00
Air Blank	0.000	13:00
Control Test	0.080	13:00
Air Blank	0.000	13:01
Subject Sample # 1	0.000	13:02
Breath Volume	3.679	13:02
Air Blank	0.000	13:02
Air Blank	0.000	13:04
Subject Sample # 2	0.000	13:05
Breath Volume	3.667	13:05
Air Blank	0.000	13:05
Control Test	0.079	13:05
Air Blank	0.000	13:06
Diagnostics Check	OK	13:06

Data Download Date/Time: 04/22/2019 09:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBB,CATHERINE,D
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: JONES,S
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 04/01/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: WASHINGTON

DATE/TIME OF TEST: 04/15/2019 23:36:55

Diagnostics Check	OK	23:34	
Air Blank	0.000	23:34	
Control Test	0.080	23:34	
Air Blank	0.000	23:35	
Subject Sample	# 1	23:35	Subject Test Refused
Breath Volume	0.000	23:35	
Air Blank	0.000	23:36	
Control Test	0.079	23:36	
Air Blank	0.000	23:36	
Diagnostics Check	OK	23:36	

Data Download Date/Time: 04/22/2019 09:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATHIS,A,J
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: REYNOLDS,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 04/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: ALBERTIE

DATE/TIME OF TEST: 04/16/2019 01:16:49

Diagnostics Check	OK	01:14	
Air Blank	0.000	01:14	
Control Test	0.080	01:14	
Air Blank	0.000	01:15	
Subject Sample	# 1	01:15	Subject Test Refused
Breath Volume	0.000	01:15	
Air Blank	0.000	01:15	
Control Test	0.080	01:16	
Air Blank	0.000	01:16	
Diagnostics Check	OK	01:16	

Data Download Date/Time: 04/22/2019 09:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,E,E
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HOLLAND,B
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 04/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: ROBBINS

DATE/TIME OF TEST: 04/18/2019 03:59:05

Diagnostics Check	OK	03:52
Air Blank	0.000	03:52
Control Test	0.080	03:53
Air Blank	0.000	03:53
Subject Sample # 1	0.126	03:54
Breath Volume	3.222	03:54
Air Blank	0.000	03:54
Air Blank	0.000	03:56
Subject Sample # 2	0.128	03:57
Breath Volume	2.628	03:57
Air Blank	0.000	03:58
Control Test	0.080	03:58
Air Blank	0.000	03:59
Diagnostics Check	OK	03:59

Data Download Date/Time: 04/22/2019 09:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,P,S
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HILLIARD,N
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 04/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 13:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:20

SUBJECT'S LAST NAME: MCNULTY

DATE/TIME OF TEST: 04/21/2019 15:01:11

Diagnostics Check	OK	14:55
Air Blank	0.000	14:55
Control Test	0.080	14:55
Air Blank	0.000	14:56
Subject Sample # 1	0.000	14:56
Breath Volume	1.726	14:56
Air Blank	0.000	14:57
Air Blank	0.000	14:59
Subject Sample # 2	0.000	14:59
Breath Volume	1.691	14:59
Air Blank	0.000	15:00
Control Test	0.081	15:00
Air Blank	0.000	15:01
Diagnostics Check	OK	15:01

Data Download Date/Time: 04/22/2019 09:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIGGINBOTHAM,B,K
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: SHEFFIELD,D
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: DEAN

DATE/TIME OF TEST: 04/21/2019 21:37:48

Diagnostics Check	OK	21:31
Air Blank	0.000	21:31
Control Test	0.079	21:32
Air Blank	0.000	21:32
Subject Sample # 1	0.228	21:33
Breath Volume	1.722	21:33
Air Blank	0.000	21:33
Air Blank	0.000	21:35
Subject Sample # 2	0.235	21:36
Breath Volume	1.449	21:36
Air Blank	0.000	21:36
Control Test	0.079	21:37
Air Blank	0.000	21:37
Diagnostics Check	OK	21:37

Data Download Date/Time: 04/22/2019 09:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,D,R
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: GOETHE,R
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 04/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: REED

DATE/TIME OF TEST: 04/21/2019 22:15:35

Diagnostics Check	OK	22:12	
Air Blank	0.000	22:13	
Control Test	0.081	22:13	
Air Blank	0.000	22:14	
Subject Sample	# 1	22:14	Subject Test Refused
Breath Volume	0.000	22:14	
Air Blank	0.000	22:14	
Control Test	0.081	22:15	
Air Blank	0.000	22:15	
Diagnostics Check	OK	22:15	

Data Download Date/Time: 04/22/2019 09:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEDDERLY,G,A
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: THOMASON,P
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:43

SUBJECT'S LAST NAME: ANLAS

DATE/TIME OF TEST: 04/26/2019 18:11:41

Diagnostics Check	OK	18:05
Air Blank	0.000	18:05
Control Test	0.082	18:05
Air Blank	0.000	18:06
Subject Sample # 1	0.000	18:07
Breath Volume	4.757	18:07
Air Blank	0.000	18:07
Air Blank	0.000	18:09
Subject Sample # 2	0.000	18:10
Breath Volume	1.789	18:10
Air Blank	0.000	18:10
Control Test	0.081	18:11
Air Blank	0.000	18:11
Diagnostics Check	OK	18:11

Data Download Date/Time: 05/28/2019 11:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KUTZ,T,E
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 04/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: LANGFORD

DATE/TIME OF TEST: 04/26/2019 22:57:57

Diagnostics Check	OK	22:51
Air Blank	0.000	22:52
Control Test	0.081	22:52
Air Blank	0.000	22:52
Subject Sample # 1	0.025	22:53
Breath Volume	1.738	22:53
Air Blank	0.000	22:54
Air Blank	0.000	22:56
Subject Sample # 2	0.025	22:56
Breath Volume	1.429	22:56
Air Blank	0.000	22:57
Control Test	0.080	22:57
Air Blank	0.000	22:57
Diagnostics Check	OK	22:57

Data Download Date/Time: 05/28/2019 11:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,E,E
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 04/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:51

SUBJECT'S LAST NAME: VANLENNEP

DATE/TIME OF TEST: 04/28/2019 00:21:15

Diagnostics Check	OK	00:14
Air Blank	0.000	00:14
Control Test	0.081	00:14
Air Blank	0.000	00:15
Subject Sample # 1	0.187	00:16
Breath Volume	2.324	00:16
Air Blank	0.000	00:17
Air Blank	0.000	00:19
Subject Sample # 2	0.195	00:19
Breath Volume	1.589	00:19
Air Blank	0.000	00:20
Control Test	0.081	00:20
Air Blank	0.000	00:21
Diagnostics Check	OK	00:21

Data Download Date/Time: 05/28/2019 11:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCracken,J,D
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 04/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: BEAL

DATE/TIME OF TEST: 04/28/2019 03:36:29

Diagnostics Check	OK	03:30
Air Blank	0.000	03:30
Control Test	0.081	03:31
Air Blank	0.000	03:31
Subject Sample # 1	0.194	03:32
Breath Volume	2.734	03:32
Air Blank	0.000	03:32
Air Blank	0.000	03:34
Subject Sample # 2	0.187	03:34
Breath Volume	3.042	03:34
Air Blank	0.000	03:35
Control Test	0.081	03:35
Air Blank	0.000	03:36
Diagnostics Check	OK	03:36

Data Download Date/Time: 05/28/2019 11:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCracken,J,D
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 04/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: HELM

DATE/TIME OF TEST: 04/28/2019 04:58:32

Diagnostics Check	OK	04:51
Air Blank	0.000	04:51
Control Test	0.081	04:51
Air Blank	0.000	04:52
Subject Sample # 1	0.198	04:54
Breath Volume	1.882	04:54
Air Blank	0.000	04:54
Air Blank	0.000	04:56
Subject Sample # 2	0.191	04:57
Breath Volume	2.234	04:57
Air Blank	0.000	04:57
Control Test	0.080	04:57
Air Blank	0.000	04:58
Diagnostics Check	OK	04:58

Data Download Date/Time: 05/28/2019 11:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEDDERLY,G,A
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HALL,J
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 04/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:41

SUBJECT'S LAST NAME: HARDY

DATE/TIME OF TEST: 04/28/2019 17:09:31

Diagnostics Check	OK	17:02
Air Blank	0.000	17:02
Control Test	0.080	17:02
Air Blank	0.000	17:03
Subject Sample # 1	0.187	17:04
Breath Volume	1.335	17:04
Air Blank	0.000	17:05
Air Blank	0.000	17:07
Subject Sample # 2	0.190	17:08
Breath Volume	1.921	17:08
Air Blank	0.000	17:08
Control Test	0.079	17:08
Air Blank	0.000	17:09
Diagnostics Check	OK	17:09

Data Download Date/Time: 05/28/2019 11:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRES,J,L
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 04/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: PERKINS

DATE/TIME OF TEST: 04/29/2019 02:06:16

Diagnostics Check	OK	01:59
Air Blank	0.000	01:59
Control Test	0.081	02:00
Air Blank	0.000	02:00
Subject Sample # 1	0.251	02:01
Breath Volume	1.550	02:01
Air Blank	0.000	02:01
Air Blank	0.000	02:03
Subject Sample # 2	0.244	02:04
Breath Volume	1.941	02:04
Air Blank	0.000	02:05
Control Test	0.080	02:05
Air Blank	0.000	02:06
Diagnostics Check	OK	02:06

Data Download Date/Time: 05/28/2019 11:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINISH,BARRY,A
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: GARRISON,JOHN
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 03/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:16

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 04/07/2019 19:39:39

Diagnostics Check	OK	19:36	
Air Blank	0.000	19:37	
Control Test	0.080	19:37	
Air Blank	0.000	19:38	
Subject Sample	# 1	19:38	Subject Test Refused
Breath Volume	0.000	19:38	
Air Blank	0.000	19:38	
Control Test	0.080	19:39	
Air Blank	0.000	19:39	
Diagnostics Check	OK	19:39	

Data Download Date/Time: 04/25/2019 11:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARLOW,ROYCE,B
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: BECK,EMILY
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 03/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: FLORES

DATE/TIME OF TEST: 04/07/2019 22:04:04

Diagnostics Check	OK	21:57
Air Blank	0.000	21:58
Control Test	0.081	21:58
Air Blank	0.000	21:59
Subject Sample # 1	0.223	21:59
Breath Volume	2.238	21:59
Air Blank	0.000	22:00
Air Blank	0.000	22:02
Subject Sample # 2	0.222	22:02
Breath Volume	2.066	22:02
Air Blank	0.000	22:03
Control Test	0.080	22:03
Air Blank	0.000	22:03
Diagnostics Check	OK	22:04

Data Download Date/Time: 04/25/2019 11:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINISH,BARRY,A
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SMITH,RYAN
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 03/10/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 04/07/2019 22:32:09

Diagnostics Check	OK	22:25
Air Blank	0.000	22:26
Control Test	0.079	22:26
Air Blank	0.000	22:26
Subject Sample # 1	0.132	22:27
Breath Volume	2.707	22:27
Air Blank	0.000	22:28
Air Blank	0.000	22:30
Subject Sample # 2	0.134	22:30
Breath Volume	3.390	22:30
Air Blank	0.000	22:31
Control Test	0.081	22:31
Air Blank	0.000	22:32
Diagnostics Check	OK	22:32

Data Download Date/Time: 04/25/2019 11:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,REBECCA,C
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: HALL,REBECCA
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 03/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: COURT ORD
ARREST TIME: 16:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: HOOKS

DATE/TIME OF TEST: 04/08/2019 16:27:11

Diagnostics Check	OK	16:20
Air Blank	0.000	16:21
Control Test	0.080	16:21
Air Blank	0.000	16:21
Subject Sample # 1	0.000	16:22
Breath Volume	3.136	16:22
Air Blank	0.000	16:23
Air Blank	0.000	16:25
Subject Sample # 2	0.000	16:25
Breath Volume	2.515	16:25
Air Blank	0.000	16:26
Control Test	0.080	16:26
Air Blank	0.000	16:27
Diagnostics Check	OK	16:27

Data Download Date/Time: 04/25/2019 11:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGOODIAL,TRISTON,T
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: MARIN,ADRIAN
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 03/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:36

SUBJECT'S LAST NAME: SANCHEZ BENITZ

DATE/TIME OF TEST: 04/12/2019 21:15:03

Diagnostics Check	OK	21:07
Air Blank	0.000	21:07
Control Test	0.080	21:08
Air Blank	0.000	21:08
Subject Sample # 1	0.300	21:10
Breath Volume	2.082	21:10
Air Blank	0.000	21:11
Air Blank	0.000	21:12
Subject Sample # 2	0.283	21:13
Breath Volume	1.585	21:13
Air Blank	0.000	21:14
Control Test	0.078	21:14
Air Blank	0.000	21:14
Diagnostics Check	OK	21:15

Data Download Date/Time: 04/25/2019 11:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,DUSTIN,J
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: CARANZA,R
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 03/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: GALVAN

DATE/TIME OF TEST: 04/14/2019 03:31:17

Diagnostics Check	OK	03:28	
Air Blank	0.000	03:28	
Control Test	0.081	03:29	
Air Blank	0.000	03:29	
Subject Sample	# 1	03:29	Subject Test Refused
Breath Volume	0.000	03:29	
Air Blank	0.000	03:30	
Control Test	0.080	03:30	
Air Blank	0.000	03:31	
Diagnostics Check	OK	03:31	

Data Download Date/Time: 04/25/2019 11:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,DUSTIN,J
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: MARIN,L
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 03/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 04/14/2019 04:26:32

Diagnostics Check	OK	04:23	
Air Blank	0.000	04:24	
Control Test	0.081	04:24	
Air Blank	0.000	04:24	
Subject Sample	# 1	04:25	Subject Test Refused
Breath Volume	0.000	04:25	
Air Blank	0.000	04:25	
Control Test	0.080	04:25	
Air Blank	0.000	04:26	
Diagnostics Check	OK	04:26	

Data Download Date/Time: 04/25/2019 11:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINISH,BARRY,A
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SALVADOR,ELDER
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 03/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 04/17/2019 02:17:59

Diagnostics Check	OK	02:11
Air Blank	0.000	02:11
Control Test	0.080	02:12
Air Blank	0.000	02:12
Subject Sample # 1	0.251	02:13
Breath Volume	1.269	02:13
Air Blank	0.000	02:14
Air Blank	0.000	02:15
Subject Sample # 2	0.261	02:16
Breath Volume	1.289	02:16
Air Blank	0.000	02:17
Control Test	0.080	02:17
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

Data Download Date/Time: 04/25/2019 11:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GLASCO,GREGORY,S
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SALVADOR,ELDER
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 04/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 04/21/2019 02:53:41

Diagnostics Check	OK	02:50	
Air Blank	0.000	02:51	
Control Test	0.080	02:51	
Air Blank	0.000	02:52	
Subject Sample	# 1	02:52	Subject Test Refused
Breath Volume	0.000	02:52	
Air Blank	0.000	02:52	
Control Test	0.081	02:53	
Air Blank	0.000	02:53	
Diagnostics Check	OK	02:53	

Data Download Date/Time: 04/25/2019 11:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINISH,BARRY,A
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: DOUBERLEY,CHADWICK
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 04/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: VIRGILE

DATE/TIME OF TEST: 04/21/2019 03:05:50

Diagnostics Check	OK	03:02	
Air Blank	0.000	03:03	
Control Test	0.080	03:03	
Air Blank	0.000	03:04	
Subject Sample	# 1	03:04	Subject Test Refused
Breath Volume	0.000	03:04	
Air Blank	0.000	03:04	
Control Test	0.079	03:05	
Air Blank	0.000	03:05	
Diagnostics Check	OK	03:05	

Data Download Date/Time: 04/25/2019 11:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNELL,C,J
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: GARRISON,J
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 04/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 04/21/2019 03:37:33

Diagnostics Check	OK	03:30
Air Blank	0.000	03:30
Control Test	0.081	03:31
Air Blank	0.000	03:31
Subject Sample # 1	0.081	03:32
Breath Volume	1.761	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.080	03:36
Breath Volume	2.242	03:36
Air Blank	0.000	03:36
Control Test	0.081	03:36
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 04/25/2019 11:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RITENOUR, STEVEN, D
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: RITENOUR, STEVEN
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 04/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: HAYNES

DATE/TIME OF TEST: 04/27/2019 01:30:58

Diagnostics Check	OK	01:23
Air Blank	0.000	01:24
Control Test	0.080	01:24
Air Blank	0.000	01:25
Subject Sample # 1	0.189	01:26
Breath Volume	2.058	01:26
Air Blank	0.000	01:26
Air Blank	0.000	01:28
Subject Sample # 2	0.193	01:29
Breath Volume	2.308	01:29
Air Blank	0.000	01:30
Control Test	0.079	01:30
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

Data Download Date/Time: 05/21/2019 10:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELLS,JOSIE,D
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: WELLS,JOSIE
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/20/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 06:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 04/27/2019 06:03:46

Diagnostics Check	OK	06:00	
Air Blank	0.000	06:01	
Control Test	0.080	06:01	
Air Blank	0.000	06:02	
Subject Sample	# 1	06:02	Subject Test Refused
Breath Volume	0.000	06:02	
Air Blank	0.000	06:02	
Control Test	0.079	06:03	
Air Blank	0.000	06:03	
Diagnostics Check	OK	06:03	

Data Download Date/Time: 05/21/2019 10:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RITENOUR,STEVEN,D
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: RITENOUR,STEVEN
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 04/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:23

SUBJECT'S LAST NAME: SANTOS-RIOS

DATE/TIME OF TEST: 04/27/2019 06:52:19

Diagnostics Check	OK	06:45
Air Blank	0.000	06:46
Control Test	0.080	06:46
Air Blank	0.000	06:47
Subject Sample # 1	0.057	06:47
Breath Volume	1.750	06:47
Air Blank	0.000	06:48
Air Blank	0.000	06:50
Subject Sample # 2	0.059	06:50
Breath Volume	2.160	06:50
Air Blank	0.000	06:51
Control Test	0.081	06:51
Air Blank	0.000	06:52
Diagnostics Check	OK	06:52

Data Download Date/Time: 05/21/2019 10:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,DUSTIN,J
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: BECERRA,R
ARRESTING OFFICER AGENCY: Lake Placid PD
LAST AGENCY INSPECTION: 04/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:04

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 04/28/2019 05:30:58

Diagnostics Check	OK	05:24
Air Blank	0.000	05:24
Control Test	0.079	05:25
Air Blank	0.000	05:25
Subject Sample # 1	0.205	05:26
Breath Volume	2.265	05:26
Air Blank	0.000	05:27
Air Blank	0.000	05:28
Subject Sample # 2	0.197	05:29
Breath Volume	2.570	05:29
Air Blank	0.000	05:30
Control Test	0.079	05:30
Air Blank	0.000	05:30
Diagnostics Check	OK	05:30

Data Download Date/Time: 05/21/2019 10:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PECK,JOHN,R
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: PEARDON,MATTHEW
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 04/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 06:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:18

SUBJECT'S LAST NAME: MASON

DATE/TIME OF TEST: 04/28/2019 07:43:00

Diagnostics Check	OK	07:39	
Air Blank	0.000	07:40	
Control Test	0.079	07:40	
Air Blank	0.000	07:41	
Subject Sample	# 1	07:41	Subject Test Refused
Breath Volume	0.000	07:41	
Air Blank	0.000	07:42	
Control Test	0.078	07:42	
Air Blank	0.000	07:42	
Diagnostics Check	OK	07:43	

Data Download Date/Time: 05/21/2019 10:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINISH,BARRY,A
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: BECK,EMILY
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 04/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 04/30/2019 02:04:18

Diagnostics Check	OK	02:03	
Air Blank	0.000	02:03	
Control Test	0.047	02:03	RFI Detect
Air Blank	0.000	02:04	

Data Download Date/Time: 05/21/2019 10:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINISH,BARRY,A
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: BECK,EMILY
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 04/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 04/30/2019 02:14:01

Diagnostics Check	OK	02:06	
Air Blank	0.000	02:06	
Control Test	0.080	02:07	
Air Blank	0.000	02:07	
Subject Sample	# 1 0.070	02:09	
Breath Volume	2.597	02:09	
Air Blank	0.000	02:09	
Air Blank	0.000	02:11	
Subject Sample	# 2 0.068	02:12	
Breath Volume	1.703	02:12	
Air Blank	0.000	02:13	
Control Test	0.080	02:13	
Air Blank		02:13	RFI Detect
Air Blank	0.000	02:14	

Data Download Date/Time: 05/21/2019 10:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

INSTRUMENT REGISTERED TO: DAVIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAHAM,HUGH,E
BREATH TEST OPERATOR AGENCY: Davie PD
ARRESTING OFFICER: STUEHRENBERG,ERIK
ARRESTING OFFICER AGENCY: Davie PD
LAST AGENCY INSPECTION: 04/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: SALAMI

DATE/TIME OF TEST: 04/04/2019 01:01:27

Diagnostics Check	OK	00:54
Air Blank	0.000	00:55
Control Test	0.080	00:55
Air Blank	0.000	00:56
Subject Sample # 1	0.371	00:57
Breath Volume	1.535	00:57
Air Blank	0.000	00:57
Air Blank	0.000	00:59
Subject Sample # 2	0.363	00:59
Breath Volume	1.746	00:59
Air Blank	0.000	01:00
Control Test	0.077	01:00
Air Blank	0.000	01:01
Diagnostics Check	OK	01:01

Data Download Date/Time: 05/13/2019 10:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

INSTRUMENT REGISTERED TO: DAVIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STUEHRENBERG,ERIK,C
BREATH TEST OPERATOR AGENCY: Davie PD
ARRESTING OFFICER: HAINES,DAVID
ARRESTING OFFICER AGENCY: Davie PD
LAST AGENCY INSPECTION: 04/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 22:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: PANDO GONZALEZ

DATE/TIME OF TEST: 04/12/2019 22:47:35

Diagnostics Check	OK	22:35	
Air Blank	0.000	22:36	
Control Test	0.081	22:36	
Air Blank	0.000	22:37	
Subject Sample	# 1 0.145	22:40	Volume Not Met
Breath Volume	0.355	22:40	
Air Blank	0.000	22:40	
Air Blank	0.000	22:42	
Subject Sample	# 2 0.153	22:45	Volume Not Met
Breath Volume	0.753	22:45	
Air Blank	0.000	22:46	
Control Test	0.079	22:46	
Air Blank	0.000	22:47	
Diagnostics Check	OK	22:47	

Data Download Date/Time: 05/13/2019 10:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

INSTRUMENT REGISTERED TO: DAVIE PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: STUEHRENBERG,ERIK,C

BREATH TEST OPERATOR AGENCY: Davie PD

ARRESTING OFFICER: HAINES,DAVID

ARRESTING OFFICER AGENCY: Davie PD

LAST AGENCY INSPECTION: 04/02/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG827002

DRY GAS EXPIRY DATE: 09/27/2020

VIOLATION CODE: DUI

ARREST TIME: 22:14

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: PANDO GONZALEZ

DATE/TIME OF TEST: 04/12/2019 23:08:36

Diagnostics Check	OK	22:57	
Air Blank	0.000	22:57	
Control Test	0.079	22:57	
Air Blank	0.000	22:58	
Subject Sample	# 1 0.141	23:01	Volume Not Met
Breath Volume	0.703	23:01	
Air Blank	0.000	23:02	
Air Blank	0.000	23:03	
Subject Sample	# 2 0.155	23:07	Volume Not Met
Breath Volume	0.015	23:07	
Air Blank	0.000	23:07	
Control Test	0.079	23:08	
Air Blank	0.000	23:08	
Diagnostics Check	OK	23:08	

Data Download Date/Time: 05/13/2019 10:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

INSTRUMENT REGISTERED TO: DAVIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORREA,MATTHEW,C
BREATH TEST OPERATOR AGENCY: Davie PD
ARRESTING OFFICER: BURNS,CLINT
ARRESTING OFFICER AGENCY: Davie PD
LAST AGENCY INSPECTION: 04/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: PACHECO

DATE/TIME OF TEST: 04/28/2019 04:24:35

Diagnostics Check	OK	04:18
Air Blank	0.000	04:18
Control Test	0.081	04:19
Air Blank	0.000	04:19
Subject Sample # 1	0.247	04:20
Breath Volume	2.527	04:20
Air Blank	0.000	04:20
Air Blank	0.000	04:22
Subject Sample # 2	0.246	04:23
Breath Volume	2.113	04:23
Air Blank	0.000	04:23
Control Test	0.081	04:23
Air Blank	0.000	04:24
Diagnostics Check	OK	04:24

Data Download Date/Time: 05/13/2019 10:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

INSTRUMENT REGISTERED TO: DAVIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORREA,MATTHEW,C
BREATH TEST OPERATOR AGENCY: Davie PD
ARRESTING OFFICER: CORREA,MATTHEW
ARRESTING OFFICER AGENCY: Davie PD
LAST AGENCY INSPECTION: 04/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: DECKER

DATE/TIME OF TEST: 04/30/2019 04:18:52

Diagnostics Check	OK	04:12
Air Blank	0.000	04:13
Control Test	0.081	04:13
Air Blank	0.000	04:14
Subject Sample # 1	0.000	04:14
Breath Volume	1.960	04:14
Air Blank	0.000	04:14
Air Blank	0.000	04:16
Subject Sample # 2	0.000	04:17
Breath Volume	2.656	04:17
Air Blank	0.000	04:17
Control Test	0.082	04:18
Air Blank	0.000	04:18
Diagnostics Check	OK	04:18

Data Download Date/Time: 05/13/2019 10:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MULDERRIG,SEAN,F

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: MULDERRIG,SEAN

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 03/25/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17817080A3

DRY GAS EXPIRY DATE: 08/05/2019

VIOLATION CODE: DUI

ARREST TIME: 17:22

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:01

SUBJECT'S LAST NAME: RICE

DATE/TIME OF TEST: 04/08/2019 22:01:26

Diagnostics Check	OK	21:55
Air Blank	0.000	21:55
Control Test	0.079	21:55
Air Blank	0.000	21:56
Subject Sample # 1	0.000	21:56
Breath Volume	4.777	21:56
Air Blank	0.000	21:57
Air Blank	0.000	21:59
Subject Sample # 2	0.000	22:00
Breath Volume	4.281	22:00
Air Blank	0.000	22:00
Control Test	0.080	22:00
Air Blank	0.000	22:01
Diagnostics Check	OK	22:01

Data Download Date/Time: 04/29/2019 13:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MULDERRIG,SEAN,F

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: MULDERRIG,SEAN

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 03/25/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17817080A3

DRY GAS EXPIRY DATE: 08/05/2019

VIOLATION CODE: DUI

ARREST TIME: 23:29

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: SAINT-LOUIS

DATE/TIME OF TEST: 04/22/2019 00:55:00

Diagnostics Check	OK	00:48
Air Blank	0.000	00:49
Control Test	0.080	00:49
Air Blank	0.000	00:50
Subject Sample # 1	0.000	00:50
Breath Volume	2.441	00:50
Air Blank	0.000	00:51
Air Blank	0.000	00:53
Subject Sample # 2	0.000	00:53
Breath Volume	2.535	00:53
Air Blank	0.000	00:54
Control Test	0.081	00:54
Air Blank	0.000	00:54
Diagnostics Check	OK	00:55

Data Download Date/Time: 04/29/2019 13:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MULDERRIG,SEAN,F

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: MULDERRIG,SEAN

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 03/25/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17817080A3

DRY GAS EXPIRY DATE: 08/05/2019

VIOLATION CODE: DUI

ARREST TIME: 16:03

GENDER: Female

OBSERVATION PERIOD BEGAN: 16:22

SUBJECT'S LAST NAME: HICKS

DATE/TIME OF TEST: 04/22/2019 16:53:05

Diagnostics Check	OK	16:45
Air Blank	0.000	16:46
Control Test	0.080	16:46
Air Blank	0.000	16:47
Subject Sample # 1	0.249	16:48
Breath Volume	1.738	16:48
Air Blank	0.000	16:48
Air Blank	0.000	16:50
Subject Sample # 2	0.256	16:51
Breath Volume	2.082	16:51
Air Blank	0.000	16:52
Control Test	0.079	16:52
Air Blank	0.000	16:53
Diagnostics Check	OK	16:53

Data Download Date/Time: 04/29/2019 13:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JASON,E
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PAETSCH,JOSH
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 03/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:56

SUBJECT'S LAST NAME: FISHER

DATE/TIME OF TEST: 04/03/2019 05:36:39

Diagnostics Check	OK	05:26
Air Blank	0.000	05:26
Control Test	0.078	05:27
Air Blank	0.000	05:27
Subject Sample # 1	0.190	05:28
Breath Volume	2.332	05:28
Air Blank	0.000	05:29
Air Blank	0.000	05:30
Subject Sample # 2	0.168	05:31
Breath Volume	2.347	05:31
Air Blank	0.000	05:32
Air Blank	0.000	05:34
Subject Sample # 3	0.178	05:35
Breath Volume	2.753	05:35
Air Blank	0.000	05:35
Control Test	0.078	05:36
Air Blank	0.000	05:36
Diagnostics Check	OK	05:36

Data Download Date/Time: 04/08/2019 14:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAETSCH,JOSH,E
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PAETSCH,JOSH
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 04/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: VAN VOORHIS

DATE/TIME OF TEST: 04/11/2019 22:58:47

Diagnostics Check	OK	22:55	
Air Blank	0.000	22:56	
Control Test	0.081	22:56	
Air Blank	0.000	22:57	
Subject Sample	# 1	22:57	Subject Test Refused
Breath Volume	0.000	22:57	
Air Blank	0.000	22:57	
Control Test	0.081	22:58	
Air Blank	0.000	22:58	
Diagnostics Check	OK	22:58	

Data Download Date/Time: 05/07/2019 13:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001062

INSTRUMENT REGISTERED TO: TREASURE ISLAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: INNES,JOSHUA,L
BREATH TEST OPERATOR AGENCY: Treasure Island PD
ARRESTING OFFICER: LAMOUREUX,MATT
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 03/27/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: IVAN

DATE/TIME OF TEST: 04/06/2019 23:24:29

Diagnostics Check	OK	23:20	
Air Blank	0.000	23:21	
Control Test	0.080	23:21	
Air Blank	0.000	23:21	
Subject Sample	# 1	23:24	Sequence Aborted
Breath Volume	0.000	23:24	
Air Blank	0.000	23:24	

Data Download Date/Time: 04/25/2019 10:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001062

INSTRUMENT REGISTERED TO: TREASURE ISLAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: INNES,JOSHUA,L
BREATH TEST OPERATOR AGENCY: Treasure Island PD
ARRESTING OFFICER: LAMOUREUX,MATT
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 03/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: IVAN

DATE/TIME OF TEST: 04/06/2019 23:32:07

Diagnostics Check	OK	23:29	
Air Blank	0.000	23:29	
Control Test	0.080	23:30	
Air Blank	0.000	23:30	
Subject Sample	# 1	23:30	Subject Test Refused
Breath Volume	0.000	23:30	
Air Blank	0.000	23:31	
Control Test	0.080	23:31	
Air Blank	0.000	23:32	
Diagnostics Check	OK	23:32	

Data Download Date/Time: 04/25/2019 10:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAZIME,OFC. A.,
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: HAZIME,OFC. A.
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: HENEVELD

DATE/TIME OF TEST: 04/01/2019 03:06:17

Diagnostics Check	OK	03:00
Air Blank	0.000	03:00
Control Test	0.079	03:00
Air Blank	0.000	03:01
Subject Sample # 1	0.068	03:01
Breath Volume	2.011	03:01
Air Blank	0.000	03:02
Air Blank	0.000	03:04
Subject Sample # 2	0.066	03:04
Breath Volume	2.234	03:04
Air Blank	0.000	03:05
Control Test	0.078	03:05
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

Data Download Date/Time: 05/02/2019 21:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAZIME,OFC.A.,
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: HAZIME,OFC.A.
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 04/13/2019 03:28:26

Diagnostics Check	OK	03:22
Air Blank	0.000	03:22
Control Test	0.079	03:22
Air Blank	0.000	03:23
Subject Sample # 1	0.213	03:23
Breath Volume	3.230	03:23
Air Blank	0.000	03:24
Air Blank	0.000	03:26
Subject Sample # 2	0.193	03:26
Breath Volume	2.433	03:26
Air Blank	0.000	03:27
Control Test	0.080	03:27
Air Blank	0.000	03:28
Diagnostics Check	OK	03:28

Data Download Date/Time: 05/02/2019 21:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAZIME,OFC. A.,
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: HARTMANN,OFC. H.
ARRESTING OFFICER AGENCY: Port Richey PD
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 04/14/2019 02:37:14

Diagnostics Check	OK	02:30
Air Blank	0.000	02:31
Control Test	0.078	02:31
Air Blank	0.000	02:32
Subject Sample # 1	0.183	02:32
Breath Volume	1.628	02:32
Air Blank	0.000	02:33
Air Blank	0.000	02:35
Subject Sample # 2	0.185	02:35
Breath Volume	1.789	02:35
Air Blank	0.000	02:36
Control Test	0.079	02:36
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

Data Download Date/Time: 05/02/2019 21:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAZIME,OFC. A.,
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: HAZIME,OFC. A.
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: SAWYER

DATE/TIME OF TEST: 04/24/2019 03:26:04

Diagnostics Check	OK	03:19
Air Blank	0.000	03:19
Control Test	0.078	03:20
Air Blank	0.000	03:20
Subject Sample # 1	0.238	03:21
Breath Volume	2.425	03:21
Air Blank	0.000	03:21
Air Blank	0.000	03:23
Subject Sample # 2	0.230	03:24
Breath Volume	1.625	03:24
Air Blank	0.000	03:25
Control Test	0.079	03:25
Air Blank	0.000	03:25
Diagnostics Check	OK	03:26

Data Download Date/Time: 05/02/2019 21:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAZIME,OFC. A.,
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: HAZIME,OFC. A.
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 04/28/2019 02:14:29

Diagnostics Check	OK	02:06
Air Blank	0.000	02:07
Control Test	0.079	02:07
Air Blank	0.000	02:08
Subject Sample # 1	0.079	02:09
Breath Volume	1.242	02:09
Air Blank	0.000	02:10
Air Blank	0.000	02:12
Subject Sample # 2	0.077	02:12
Breath Volume	1.476	02:12
Air Blank	0.000	02:13
Control Test	0.079	02:13
Air Blank	0.000	02:14
Diagnostics Check	OK	02:14

Data Download Date/Time: 05/02/2019 21:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPATZ,JOHN,M
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: ROSE,BRITTNEY
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: HERD

DATE/TIME OF TEST: 04/02/2019 19:50:55

Diagnostics Check	OK	19:48	
Air Blank	0.000	19:48	
Control Test	0.080	19:48	
Air Blank	0.000	19:49	
Subject Sample	# 1	19:49	Subject Test Refused
Breath Volume	0.000	19:49	
Air Blank	0.000	19:50	
Control Test	0.079	19:50	
Air Blank	0.000	19:50	
Diagnostics Check	OK	19:50	

Data Download Date/Time: 05/06/2019 11:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPATZ,JOHN,M
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: GEER,M
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:40

SUBJECT'S LAST NAME: SZAL

DATE/TIME OF TEST: 04/20/2019 18:16:20

Diagnostics Check	OK	18:10
Air Blank	0.000	18:10
Control Test	0.081	18:10
Air Blank	0.000	18:11
Subject Sample # 1	0.272	18:11
Breath Volume	1.660	18:11
Air Blank	0.000	18:12
Air Blank	0.000	18:14
Subject Sample # 2	0.264	18:14
Breath Volume	2.574	18:14
Air Blank	0.000	18:15
Control Test	0.078	18:15
Air Blank	0.000	18:16
Diagnostics Check	OK	18:16

Data Download Date/Time: 05/06/2019 11:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RONDOS,ROBERT,E
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: RONDOS,ROBERT
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:53

SUBJECT'S LAST NAME: NEGRON

DATE/TIME OF TEST: 04/22/2019 19:23:35

Diagnostics Check	OK	19:16
Air Blank	0.000	19:17
Control Test	0.080	19:17
Air Blank	0.000	19:18
Subject Sample # 1	0.263	19:18
Breath Volume	1.394	19:18
Air Blank	0.000	19:19
Air Blank	0.000	19:21
Subject Sample # 2	0.255	19:22
Breath Volume	1.824	19:22
Air Blank	0.000	19:22
Control Test	0.078	19:22
Air Blank	0.000	19:23
Diagnostics Check	OK	19:23

Data Download Date/Time: 05/06/2019 11:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GRUNDMANN,B
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: WHITT-MORGAN

DATE/TIME OF TEST: 04/01/2019 00:15:31

Diagnostics Check	OK	00:11	
Air Blank	0.000	00:11	
Control Test	0.081	00:11	
Air Blank	0.000	00:12	
Subject Sample	# 1	00:14	Subject Test Refused
Breath Volume	0.000	00:14	
Air Blank	0.000	00:14	
Control Test	0.080	00:14	
Air Blank	0.000	00:15	
Diagnostics Check	OK	00:15	

Data Download Date/Time: 04/18/2019 04:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: BOTKO

DATE/TIME OF TEST: 04/01/2019 02:38:16

Diagnostics Check	OK	02:29
Air Blank	0.000	02:29
Control Test	0.082	02:30
Air Blank	0.000	02:30
Subject Sample # 1	0.160	02:31
Breath Volume	1.335	02:31
Air Blank	0.000	02:32
Air Blank	0.000	02:34
Subject Sample # 2	0.162	02:36
Breath Volume	1.839	02:36
Air Blank	0.000	02:37
Control Test	0.080	02:37
Air Blank	0.000	02:38
Diagnostics Check	OK	02:38

Data Download Date/Time: 04/18/2019 04:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,NELSON,X
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CARLSON,SEAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: FELIX

DATE/TIME OF TEST: 04/03/2019 01:58:23

Diagnostics Check	OK	01:49
Air Blank	0.000	01:50
Control Test	0.082	01:50
Air Blank	0.000	01:50
Subject Sample # 1	0.220	01:53
Breath Volume	1.949	01:53
Air Blank	0.000	01:54
Air Blank	0.000	01:56
Subject Sample # 2	0.210	01:56
Breath Volume	2.859	01:56
Air Blank	0.000	01:57
Control Test	0.081	01:57
Air Blank	0.000	01:58
Diagnostics Check	OK	01:58

Data Download Date/Time: 04/18/2019 04:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOOD,COREY,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SIEM,DIMAS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 07:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:38

SUBJECT'S LAST NAME: SANTIAGO-RAMIREZ

DATE/TIME OF TEST: 04/03/2019 10:10:45

Diagnostics Check	OK	10:05	
Air Blank	0.000	10:06	
Control Test	0.082	10:06	
Air Blank	0.000	10:06	
Subject Sample	# 1	10:09	Subject Test Refused
Breath Volume	0.000	10:09	
Air Blank	0.000	10:09	
Control Test	0.082	10:10	
Air Blank	0.000	10:10	
Diagnostics Check	OK	10:10	

Data Download Date/Time: 04/18/2019 04:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: DIOMETRE

DATE/TIME OF TEST: 04/04/2019 03:23:51

Diagnostics Check	OK	03:20	
Air Blank	0.000	03:20	
Control Test	0.082	03:20	
Air Blank	0.000	03:21	
Subject Sample	# 1	03:22	Subject Test Refused
Breath Volume	0.000	03:22	
Air Blank	0.000	03:22	
Control Test	0.082	03:23	
Air Blank	0.000	03:23	
Diagnostics Check	OK	03:23	

Data Download Date/Time: 04/18/2019 04:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICOTTILLI, JOSHUA, A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI, JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: FAGAN

DATE/TIME OF TEST: 04/04/2019 04:09:04

Diagnostics Check	OK	04:04	
Air Blank	0.000	04:04	
Control Test	0.082	04:05	
Air Blank	0.000	04:05	
Subject Sample	# 1	04:07	Subject Test Refused
Breath Volume	0.000	04:07	
Air Blank	0.000	04:08	
Control Test	0.082	04:08	
Air Blank	0.000	04:08	
Diagnostics Check	OK	04:09	

Data Download Date/Time: 04/18/2019 04:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEELER,BRYAN,C
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LAMBIRD,DAVID
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 07:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:35

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 04/04/2019 09:05:05

Diagnostics Check	OK	09:04	
Air Blank		09:04	RFI Detect
Air Blank	0.000	09:05	

Data Download Date/Time: 04/18/2019 04:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEELER,BRYAN,C
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LAMBIRD,DAVID
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 07:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:35

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 04/04/2019 09:15:25

Diagnostics Check	OK	09:09
Air Blank	0.000	09:09
Control Test	0.082	09:10
Air Blank	0.000	09:10
Subject Sample # 1	0.000	09:11
Breath Volume	2.878	09:11
Air Blank	0.000	09:11
Air Blank	0.000	09:13
Subject Sample # 2	0.000	09:14
Breath Volume	3.035	09:14
Air Blank	0.000	09:14
Control Test	0.082	09:14
Air Blank	0.000	09:15
Diagnostics Check	OK	09:15

Data Download Date/Time: 04/18/2019 04:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: CLEMONS

DATE/TIME OF TEST: 04/05/2019 02:17:55

Diagnostics Check	OK	02:11
Air Blank	0.000	02:12
Control Test	0.082	02:12
Air Blank	0.000	02:13
Subject Sample # 1	0.000	02:13
Breath Volume	2.757	02:13
Air Blank	0.000	02:14
Air Blank	0.000	02:16
Subject Sample # 2	0.000	02:16
Breath Volume	3.343	02:16
Air Blank	0.000	02:17
Control Test	0.082	02:17
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

Data Download Date/Time: 04/18/2019 04:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RANISZEWSKI,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 04/05/2019 22:34:26

Diagnostics Check	OK	22:27
Air Blank	0.000	22:28
Control Test	0.082	22:28
Air Blank	0.000	22:29
Subject Sample # 1	0.000	22:30
Breath Volume	1.683	22:30
Air Blank	0.000	22:30
Air Blank	0.000	22:32
Subject Sample # 2	0.000	22:33
Breath Volume	1.234	22:33
Air Blank	0.000	22:33
Control Test	0.081	22:33
Air Blank	0.000	22:34
Diagnostics Check	OK	22:34

Data Download Date/Time: 04/18/2019 04:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: FURMAN

DATE/TIME OF TEST: 04/05/2019 23:09:57

Diagnostics Check	OK	23:02
Air Blank	0.000	23:03
Control Test	0.082	23:03
Air Blank	0.000	23:03
Subject Sample # 1	0.168	23:04
Breath Volume	1.378	23:04
Air Blank	0.000	23:05
Air Blank	0.000	23:07
Subject Sample # 2	0.171	23:08
Breath Volume	1.421	23:08
Air Blank	0.000	23:09
Control Test	0.081	23:09
Air Blank	0.000	23:09
Diagnostics Check	OK	23:09

Data Download Date/Time: 04/18/2019 04:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CASTELEIRO,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: SUMPTER

DATE/TIME OF TEST: 04/06/2019 00:12:30

Diagnostics Check	OK	00:00	
Air Blank	0.000	00:00	
Control Test	0.081	00:00	
Air Blank	0.000	00:01	
Subject Sample	# 1 0.292	00:02	
Breath Volume	2.343	00:02	
Air Blank	0.000	00:03	
Air Blank	0.000	00:05	
Subject Sample	# 2	00:08	No Sample Provided
Breath Volume	0.000	00:08	
Air Blank	0.000	00:08	
Air Blank	0.000	00:10	
Subject Sample	# 3	00:11	Subject Test Refused
Breath Volume	0.000	00:11	
Air Blank	0.000	00:11	
Control Test	0.080	00:11	
Air Blank	0.000	00:12	
Diagnostics Check	OK	00:12	

Data Download Date/Time: 04/18/2019 04:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RANISZEWSKI,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: CARRION HENRIQUEZ

DATE/TIME OF TEST: 04/06/2019 02:07:12

Diagnostics Check	OK	02:04	
Air Blank	0.000	02:04	
Control Test	0.082	02:05	
Air Blank	0.000	02:05	
Subject Sample	# 1	02:05	Subject Test Refused
Breath Volume	0.000	02:05	
Air Blank	0.000	02:06	
Control Test	0.082	02:06	
Air Blank	0.000	02:07	
Diagnostics Check	OK	02:07	

Data Download Date/Time: 04/18/2019 04:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 04/06/2019 03:14:07

Diagnostics Check	OK	03:11	
Air Blank	0.000	03:11	
Control Test	0.081	03:12	
Air Blank	0.000	03:12	
Subject Sample	# 1	03:12	Subject Test Refused
Breath Volume	0.000	03:12	
Air Blank	0.000	03:13	
Control Test	0.081	03:13	
Air Blank	0.000	03:14	
Diagnostics Check	OK	03:14	

Data Download Date/Time: 04/18/2019 04:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: VAZQUEZ BAEZ

DATE/TIME OF TEST: 04/06/2019 04:07:53

Diagnostics Check	OK	04:01
Air Blank	0.000	04:02
Control Test	0.082	04:02
Air Blank	0.000	04:02
Subject Sample # 1	0.037	04:03
Breath Volume	2.421	04:03
Air Blank	0.000	04:04
Air Blank	0.000	04:06
Subject Sample # 2	0.031	04:06
Breath Volume	1.105	04:06
Air Blank	0.000	04:06
Control Test	0.082	04:07
Air Blank	0.000	04:07
Diagnostics Check	OK	04:07

Data Download Date/Time: 04/18/2019 04:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PEREZ,Y
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: CALDWELL

DATE/TIME OF TEST: 04/06/2019 23:13:52

Diagnostics Check	OK	23:06
Air Blank	0.000	23:06
Control Test	0.082	23:06
Air Blank	0.000	23:07
Subject Sample # 1	0.229	23:09
Breath Volume	1.402	23:09
Air Blank	0.000	23:09
Air Blank	0.000	23:11
Subject Sample # 2	0.227	23:12
Breath Volume	1.414	23:12
Air Blank	0.000	23:12
Control Test	0.081	23:13
Air Blank	0.000	23:13
Diagnostics Check	OK	23:13

Data Download Date/Time: 04/18/2019 04:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RANISZEWSKI,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: WALKER-DOESCHER

DATE/TIME OF TEST: 04/07/2019 01:06:13

Diagnostics Check	OK	00:54	
Air Blank	0.000	00:55	
Control Test	0.082	00:55	
Air Blank	0.000	00:56	
Subject Sample	# 1	00:59	No Sample Provided
Breath Volume	0.000	00:59	
Air Blank	0.000	00:59	
Air Blank	0.000	01:01	
Subject Sample	# 2	01:04	No Sample Provided
Breath Volume	0.000	01:04	
Air Blank	0.000	01:05	
Control Test	0.081	01:05	
Air Blank	0.000	01:06	
Diagnostics Check	OK	01:06	

Data Download Date/Time: 04/18/2019 04:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: DOYON

DATE/TIME OF TEST: 04/07/2019 02:42:52

Diagnostics Check	OK	02:35
Air Blank	0.000	02:36
Control Test	0.082	02:36
Air Blank	0.000	02:37
Subject Sample # 1	0.095	02:38
Breath Volume	1.933	02:38
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample # 2	0.095	02:41
Breath Volume	1.886	02:41
Air Blank	0.000	02:41
Control Test	0.081	02:42
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 04/18/2019 04:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE, ELIZABETH,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE, ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: STRATON

DATE/TIME OF TEST: 04/09/2019 00:24:42

Diagnostics Check	OK	00:17
Air Blank	0.000	00:17
Control Test	0.082	00:18
Air Blank	0.000	00:18
Subject Sample # 1	0.112	00:19
Breath Volume	1.246	00:19
Air Blank	0.000	00:20
Air Blank	0.000	00:22
Subject Sample # 2	0.113	00:23
Breath Volume	1.613	00:23
Air Blank	0.000	00:23
Control Test	0.082	00:24
Air Blank	0.000	00:24
Diagnostics Check	OK	00:24

Data Download Date/Time: 04/18/2019 04:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BREGON,JAMES,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BREGON,JAMES
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: PAGLIUCA

DATE/TIME OF TEST: 04/09/2019 04:10:00

Diagnostics Check	OK	04:03
Air Blank	0.000	04:03
Control Test	0.082	04:04
Air Blank	0.000	04:04
Subject Sample # 1	0.077	04:05
Breath Volume	1.949	04:05
Air Blank	0.000	04:06
Air Blank	0.000	04:07
Subject Sample # 2	0.078	04:08
Breath Volume	1.890	04:08
Air Blank	0.000	04:09
Control Test	0.082	04:09
Air Blank	0.000	04:09
Diagnostics Check	OK	04:10

Data Download Date/Time: 04/18/2019 04:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FOSTER,ALEXANDER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: POMEROY

DATE/TIME OF TEST: 04/09/2019 05:08:14

Diagnostics Check	OK	05:02
Air Blank	0.000	05:02
Control Test	0.081	05:02
Air Blank	0.000	05:03
Subject Sample # 1	0.266	05:03
Breath Volume	2.503	05:03
Air Blank	0.000	05:04
Air Blank	0.000	05:06
Subject Sample # 2	0.263	05:06
Breath Volume	1.988	05:06
Air Blank	0.000	05:07
Control Test	0.080	05:07
Air Blank	0.000	05:08
Diagnostics Check	OK	05:08

Data Download Date/Time: 04/18/2019 04:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUDTKE,CHRISTOPHER,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LUDTKE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:52

SUBJECT'S LAST NAME: SMADES

DATE/TIME OF TEST: 04/09/2019 19:21:31

Diagnostics Check	OK	19:19	
Air Blank	0.000	19:20	
Control Test	0.082	19:20	
Air Blank	0.000	19:20	
		19:21	RFI Detect
Air Blank	0.000	19:21	

Data Download Date/Time: 04/18/2019 04:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUDTKE,CHRISTOPHER,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LUDTKE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:52

SUBJECT'S LAST NAME: SMADES

DATE/TIME OF TEST: 04/09/2019 19:26:28

Diagnostics Check	OK	19:23	
Air Blank	0.000	19:24	
Control Test	0.081	19:24	
Air Blank	0.000	19:24	
Subject Sample	# 1	19:25	Subject Test Refused
Breath Volume	0.000	19:25	
Air Blank	0.000	19:25	
Control Test	0.081	19:25	
Air Blank	0.000	19:26	
Diagnostics Check	OK	19:26	

Data Download Date/Time: 04/18/2019 04:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: CHASE

DATE/TIME OF TEST: 04/10/2019 00:07:30

Diagnostics Check	OK	00:01
Air Blank	0.000	00:01
Control Test	0.081	00:01
Air Blank	0.000	00:02
Subject Sample # 1	0.347	00:02
Breath Volume	2.746	00:02
Air Blank	0.000	00:03
Air Blank	0.000	00:05
Subject Sample # 2	0.330	00:05
Breath Volume	3.300	00:05
Air Blank	0.000	00:06
Control Test	0.081	00:06
Air Blank	0.000	00:07
Diagnostics Check	OK	00:07

Data Download Date/Time: 04/18/2019 04:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: REIDY,T
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/20/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: LENIART

DATE/TIME OF TEST: 04/11/2019 02:13:02

Diagnostics Check	OK	02:01	
Air Blank	0.000	02:02	
Control Test	0.082	02:02	
Air Blank	0.000	02:02	
Subject Sample	# 1	02:06	No Sample Provided
Breath Volume	0.000	02:06	
Air Blank	0.000	02:06	
Air Blank	0.000	02:08	
Subject Sample	# 2	02:11	No Sample Provided
Breath Volume	0.000	02:11	
Air Blank	0.000	02:12	
Control Test	0.082	02:12	
Air Blank	0.000	02:12	
Diagnostics Check	OK	02:13	

Data Download Date/Time: 04/18/2019 04:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: REIDY,T
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: LENIART

DATE/TIME OF TEST: 04/11/2019 02:23:32

Diagnostics Check	OK	02:15
Air Blank	0.000	02:15
Control Test	0.082	02:15
Air Blank	0.000	02:16
Subject Sample # 1	0.147	02:18
Breath Volume	1.394	02:18
Air Blank	0.000	02:18
Air Blank	0.000	02:20
Subject Sample # 2	0.162	02:21
Breath Volume	2.312	02:21
Air Blank	0.000	02:22
Control Test	0.081	02:22
Air Blank	0.000	02:23
Diagnostics Check	OK	02:23

Data Download Date/Time: 04/18/2019 04:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEDE,BRENNEN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WEDE,BRENNEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: HIGUEROS

DATE/TIME OF TEST: 04/11/2019 03:22:46

Diagnostics Check	OK	03:15
Air Blank	0.000	03:16
Control Test	0.082	03:16
Air Blank	0.000	03:16
Subject Sample # 1	0.016	03:17
Breath Volume	1.457	03:17
Air Blank	0.000	03:18
Air Blank	0.000	03:20
Subject Sample # 2	0.015	03:21
Breath Volume	1.402	03:21
Air Blank	0.000	03:21
Control Test	0.081	03:22
Air Blank	0.000	03:22
Diagnostics Check	OK	03:22

Data Download Date/Time: 04/18/2019 04:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOBSON,JON,B
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DOBSON,JON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 11:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:24

SUBJECT'S LAST NAME: KONDZIELLA

DATE/TIME OF TEST: 04/11/2019 13:00:11

Diagnostics Check	OK	12:50
Air Blank	0.000	12:50
Control Test	0.082	12:50
Air Blank	0.000	12:51
Subject Sample # 1	0.327	12:52
Breath Volume	4.097	12:52
Air Blank	0.000	12:52
Air Blank	0.000	12:54
Subject Sample # 2	0.304	12:55
Breath Volume	2.433	12:55
Air Blank	0.000	12:56
Air Blank	0.000	12:57
Subject Sample # 3	0.344	12:58
Breath Volume	3.804	12:58
Air Blank	0.000	12:59
Control Test	0.081	12:59
Air Blank	0.000	13:00
Diagnostics Check	OK	13:00

Data Download Date/Time: 04/18/2019 04:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:48

SUBJECT'S LAST NAME: HARDY

DATE/TIME OF TEST: 04/11/2019 19:14:52

Diagnostics Check	OK	19:08
Air Blank	0.000	19:08
Control Test	0.079	19:09
Air Blank	0.000	19:09
Subject Sample # 1	0.189	19:10
Breath Volume	1.886	19:10
Air Blank	0.000	19:10
Air Blank	0.000	19:12
Subject Sample # 2	0.187	19:13
Breath Volume	1.898	19:13
Air Blank	0.000	19:13
Control Test	0.082	19:14
Air Blank	0.000	19:14
Diagnostics Check	OK	19:14

Data Download Date/Time: 04/18/2019 04:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BREGON,JAMES,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BREGON,JAMES
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: HINDS

DATE/TIME OF TEST: 04/12/2019 01:13:04

Diagnostics Check	OK	01:06
Air Blank	0.000	01:06
Control Test	0.082	01:07
Air Blank	0.000	01:07
Subject Sample # 1	0.273	01:08
Breath Volume	1.527	01:08
Air Blank	0.000	01:09
Air Blank	0.000	01:10
Subject Sample # 2	0.269	01:11
Breath Volume	1.503	01:11
Air Blank	0.000	01:12
Control Test	0.080	01:12
Air Blank	0.000	01:12
Diagnostics Check	OK	01:13

Data Download Date/Time: 04/18/2019 04:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: COOK,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 04/12/2019 02:43:17

Diagnostics Check	OK	02:36
Air Blank	0.000	02:36
Control Test	0.081	02:36
Air Blank	0.000	02:37
Subject Sample # 1	0.201	02:38
Breath Volume	2.699	02:38
Air Blank	0.000	02:39
Air Blank	0.000	02:41
Subject Sample # 2	0.187	02:41
Breath Volume	1.984	02:41
Air Blank	0.000	02:42
Control Test	0.079	02:42
Air Blank	0.000	02:43
Diagnostics Check	OK	02:43

Data Download Date/Time: 04/18/2019 04:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEDE,BRENNEN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WEDE,BRENNEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 14:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:29

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 04/12/2019 15:56:28

Diagnostics Check	OK	15:50
Air Blank	0.000	15:50
Control Test	0.082	15:50
Air Blank	0.000	15:51
Subject Sample # 1	0.000	15:51
Breath Volume	1.308	15:51
Air Blank	0.000	15:52
Air Blank	0.000	15:54
Subject Sample # 2	0.000	15:55
Breath Volume	1.339	15:55
Air Blank	0.000	15:55
Control Test	0.081	15:55
Air Blank	0.000	15:56
Diagnostics Check	OK	15:56

Data Download Date/Time: 04/18/2019 04:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ROSADO,I
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: HIGGINS

DATE/TIME OF TEST: 04/12/2019 22:05:13

Diagnostics Check	OK	21:58
Air Blank	0.000	21:58
Control Test	0.081	21:58
Air Blank	0.000	21:59
Subject Sample # 1	0.146	22:00
Breath Volume	1.687	22:00
Air Blank	0.000	22:00
Air Blank	0.000	22:02
Subject Sample # 2	0.148	22:03
Breath Volume	1.699	22:03
Air Blank	0.000	22:04
Control Test	0.081	22:04
Air Blank	0.000	22:05
Diagnostics Check	OK	22:05

Data Download Date/Time: 04/18/2019 04:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANGE,K
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: MUNSON

DATE/TIME OF TEST: 04/12/2019 22:13:12

Diagnostics Check	OK	22:10	
Air Blank	0.000	22:10	
Control Test	0.081	22:11	
Air Blank	0.000	22:11	
Subject Sample	# 1	22:11	Subject Test Refused
Breath Volume	0.000	22:11	
Air Blank	0.000	22:12	
Control Test	0.081	22:12	
Air Blank	0.000	22:13	
Diagnostics Check	OK	22:13	

Data Download Date/Time: 04/18/2019 04:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FOSTER,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: MARGOLIS

DATE/TIME OF TEST: 04/13/2019 02:56:04

Diagnostics Check	OK	02:49
Air Blank	0.000	02:50
Control Test	0.082	02:50
Air Blank	0.000	02:51
Subject Sample # 1	0.124	02:51
Breath Volume	1.390	02:51
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample # 2	0.120	02:54
Breath Volume	2.644	02:54
Air Blank	0.000	02:55
Control Test	0.081	02:55
Air Blank	0.000	02:55
Diagnostics Check	OK	02:56

Data Download Date/Time: 04/18/2019 04:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MICHELL,R
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: WHITTAKER

DATE/TIME OF TEST: 04/13/2019 03:07:17

Diagnostics Check	OK	03:00
Air Blank	0.000	03:00
Control Test	0.081	03:01
Air Blank	0.000	03:01
Subject Sample # 1	0.176	03:02
Breath Volume	2.410	03:02
Air Blank	0.000	03:03
Air Blank	0.000	03:05
Subject Sample # 2	0.177	03:05
Breath Volume	2.691	03:05
Air Blank	0.000	03:06
Control Test	0.081	03:06
Air Blank	0.000	03:07
Diagnostics Check	OK	03:07

Data Download Date/Time: 04/18/2019 04:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, WILLIAM, T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: DIAZ CRUZ

DATE/TIME OF TEST: 04/13/2019 04:21:44

Diagnostics Check	OK	04:08	
Air Blank	0.000	04:08	
Control Test	0.082	04:09	
Air Blank	0.000	04:09	
Subject Sample	# 1 0.180	04:12	Volume Not Met
Breath Volume	1.082	04:12	
Air Blank	0.000	04:13	
Air Blank	0.000	04:15	
Subject Sample	# 2 0.184	04:16	
Breath Volume	1.171	04:16	
Air Blank	0.000	04:17	
Air Blank	0.000	04:19	
Subject Sample	# 3 0.184	04:20	
Breath Volume	2.269	04:20	
Air Blank	0.000	04:20	
Control Test	0.080	04:21	
Air Blank	0.000	04:21	
Diagnostics Check	OK	04:21	

Data Download Date/Time: 04/18/2019 04:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOOD,COREY,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CONEBY,MEGAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 06:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:30

SUBJECT'S LAST NAME: OAKS

DATE/TIME OF TEST: 04/13/2019 07:59:47

Diagnostics Check	OK	07:50
Air Blank	0.000	07:51
Control Test	0.081	07:51
Air Blank	0.000	07:52
Subject Sample # 1	0.166	07:54
Breath Volume	1.593	07:54
Air Blank	0.000	07:54
Air Blank	0.000	07:56
Subject Sample # 2	0.156	07:58
Breath Volume	2.171	07:58
Air Blank	0.000	07:58
Control Test	0.080	07:59
Air Blank	0.000	07:59
Diagnostics Check	OK	07:59

Data Download Date/Time: 04/18/2019 04:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSEN,STEPHEN,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: OREILLY,TERRENCE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:12

SUBJECT'S LAST NAME: JOHNSTON

DATE/TIME OF TEST: 04/13/2019 19:39:23

Diagnostics Check	OK	19:32
Air Blank	0.000	19:33
Control Test	0.082	19:33
Air Blank	0.000	19:34
Subject Sample # 1	0.092	19:34
Breath Volume	1.726	19:34
Air Blank	0.000	19:35
Air Blank	0.000	19:36
Subject Sample # 2	0.089	19:37
Breath Volume	1.875	19:37
Air Blank	0.000	19:38
Control Test	0.081	19:38
Air Blank	0.000	19:39
Diagnostics Check	OK	19:39

Data Download Date/Time: 04/18/2019 04:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DONOHUE,LAWRENCE,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DONOHUE,LAWRENCE
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: ARMSTRONG

DATE/TIME OF TEST: 04/13/2019 21:40:44

Diagnostics Check	OK	21:35	
Air Blank	0.000	21:35	
Control Test	0.082	21:36	
Air Blank	0.000	21:36	
Subject Sample	# 1	21:39	Subject Test Refused
Breath Volume	0.000	21:39	
Air Blank	0.000	21:39	
Control Test	0.081	21:40	
Air Blank	0.000	21:40	
Diagnostics Check	OK	21:40	

Data Download Date/Time: 04/18/2019 04:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSEN,STEPHEN,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MASON,GERALD
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: MATTHEWS

DATE/TIME OF TEST: 04/14/2019 03:21:06

Diagnostics Check	OK	03:14
Air Blank	0.000	03:15
Control Test	0.082	03:15
Air Blank	0.000	03:15
Subject Sample # 1	0.221	03:16
Breath Volume	1.683	03:16
Air Blank	0.000	03:17
Air Blank	0.000	03:18
Subject Sample # 2	0.217	03:19
Breath Volume	1.667	03:19
Air Blank	0.000	03:20
Control Test	0.081	03:20
Air Blank	0.000	03:21
Diagnostics Check	OK	03:21

Data Download Date/Time: 04/18/2019 04:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FEIDT,E
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: YANG

DATE/TIME OF TEST: 04/15/2019 01:12:57

Diagnostics Check	OK	01:05
Air Blank	0.000	01:06
Control Test	0.081	01:06
Air Blank	0.000	01:07
Subject Sample # 1	0.161	01:08
Breath Volume	1.609	01:08
Air Blank	0.000	01:09
Air Blank	0.000	01:10
Subject Sample # 2	0.173	01:11
Breath Volume	1.789	01:11
Air Blank	0.000	01:12
Control Test	0.081	01:12
Air Blank	0.000	01:12
Diagnostics Check	OK	01:12

Data Download Date/Time: 04/18/2019 04:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE,DEPUTY WILLIAM,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,DEPUTY L.
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:23

SUBJECT'S LAST NAME: MASTORAKIS

DATE/TIME OF TEST: 04/15/2019 21:52:56

Diagnostics Check	OK	21:46
Air Blank	0.000	21:46
Control Test	0.082	21:46
Air Blank	0.000	21:47
Subject Sample # 1	0.098	21:48
Breath Volume	2.425	21:48
Air Blank	0.000	21:48
Air Blank	0.000	21:50
Subject Sample # 2	0.097	21:51
Breath Volume	1.960	21:51
Air Blank	0.000	21:52
Control Test	0.081	21:52
Air Blank	0.000	21:52
Diagnostics Check	OK	21:52

Data Download Date/Time: 04/18/2019 04:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:57

SUBJECT'S LAST NAME: DUNN

DATE/TIME OF TEST: 04/16/2019 20:24:41

Diagnostics Check	OK	20:17
Air Blank	0.000	20:17
Control Test	0.082	20:18
Air Blank	0.000	20:18
Subject Sample # 1	0.120	20:19
Breath Volume	1.691	20:19
Air Blank	0.000	20:20
Air Blank	0.000	20:22
Subject Sample # 2	0.113	20:23
Breath Volume	1.683	20:23
Air Blank	0.000	20:23
Control Test	0.081	20:24
Air Blank	0.000	20:24
Diagnostics Check	OK	20:24

Data Download Date/Time: 04/18/2019 04:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERSAUD,COMPTON,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PERSAUD,COMPTON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 04/17/2019 00:14:42

Diagnostics Check	OK	00:07
Air Blank	0.000	00:08
Control Test	0.080	00:08
Air Blank	0.000	00:09
Subject Sample # 1	0.206	00:10
Breath Volume	1.542	00:10
Air Blank	0.000	00:10
Air Blank	0.000	00:12
Subject Sample # 2	0.203	00:13
Breath Volume	1.156	00:13
Air Blank	0.000	00:13
Control Test	0.080	00:14
Air Blank	0.000	00:14
Diagnostics Check	OK	00:14

Data Download Date/Time: 04/18/2019 04:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MASON,G
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:37

SUBJECT'S LAST NAME: GANIER

DATE/TIME OF TEST: 04/17/2019 22:05:34

Diagnostics Check	OK	21:58
Air Blank	0.000	21:58
Control Test	0.082	21:58
Air Blank	0.000	21:59
Subject Sample # 1	0.181	22:00
Breath Volume	1.457	22:00
Air Blank	0.000	22:01
Air Blank	0.000	22:03
Subject Sample # 2	0.184	22:04
Breath Volume	1.761	22:04
Air Blank	0.000	22:04
Control Test	0.079	22:04
Air Blank	0.000	22:05
Diagnostics Check	OK	22:05

Data Download Date/Time: 04/18/2019 04:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOOD,COREY,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MARTINEZ,MARCOS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:56

SUBJECT'S LAST NAME: RESOVSKY

DATE/TIME OF TEST: 04/18/2019 17:22:36

Diagnostics Check	OK	17:16
Air Blank	0.000	17:16
Control Test	0.080	17:17
Air Blank	0.000	17:17
Subject Sample # 1	0.253	17:18
Breath Volume	1.769	17:18
Air Blank	0.000	17:18
Air Blank	0.000	17:20
Subject Sample # 2	0.246	17:21
Breath Volume	1.656	17:21
Air Blank	0.000	17:21
Control Test	0.081	17:22
Air Blank	0.000	17:22
Diagnostics Check	OK	17:22

Data Download Date/Time: 05/11/2019 23:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUDTKE,CHRISTOPHER,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LUDTKE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 04/18/2019 23:20:16

Diagnostics Check	OK	23:17	
Air Blank	0.000	23:17	
Control Test	0.082	23:18	
Air Blank	0.000	23:18	
Subject Sample	# 1	23:18	Subject Test Refused
Breath Volume	0.000	23:18	
Air Blank	0.000	23:19	
Control Test	0.082	23:19	
Air Blank	0.000	23:20	
Diagnostics Check	OK	23:20	

Data Download Date/Time: 05/11/2019 23:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEDE,BRENNEN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WEDE,BRENNEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: CAMPILLO

DATE/TIME OF TEST: 04/19/2019 00:49:48

Diagnostics Check	OK	00:45	
Air Blank	0.000	00:46	
Control Test	0.081	00:46	
Air Blank	0.000	00:47	
Subject Sample	# 1	00:48	Subject Test Refused
Breath Volume	0.000	00:48	
Air Blank	0.000	00:48	
Control Test	0.082	00:49	
Air Blank	0.000	00:49	
Diagnostics Check	OK	00:49	

Data Download Date/Time: 05/11/2019 23:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 04/19/2019 02:31:10

Diagnostics Check	OK	02:22	
Air Blank	0.000	02:22	
Control Test	0.082	02:22	
Air Blank	0.000	02:23	
Subject Sample	# 1	02:26	No Sample Provided
Breath Volume	0.000	02:26	
Air Blank	0.000	02:26	
Air Blank	0.000	02:28	
Subject Sample	# 2	02:29	Subject Test Refused
Breath Volume	0.000	02:29	
Air Blank	0.000	02:30	
Control Test	0.082	02:30	
Air Blank	0.000	02:31	
Diagnostics Check	OK	02:31	

Data Download Date/Time: 05/11/2019 23:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEDE,BRENNEN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WEDE,BRENNEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:46

SUBJECT'S LAST NAME: TRIPLETT

DATE/TIME OF TEST: 04/19/2019 05:10:22

Diagnostics Check	OK	05:07	
Air Blank	0.000	05:08	
Control Test	0.082	05:08	
Air Blank	0.000	05:08	
Subject Sample	# 1	05:09	Subject Test Refused
Breath Volume	0.000	05:09	
Air Blank	0.000	05:09	
Control Test	0.082	05:09	
Air Blank	0.000	05:10	
Diagnostics Check	OK	05:10	

Data Download Date/Time: 05/11/2019 23:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: REIDY,T
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: CALIXTRO-RAMOS

DATE/TIME OF TEST: 04/19/2019 22:53:02

Diagnostics Check	OK	22:46
Air Blank	0.000	22:46
Control Test	0.082	22:46
Air Blank	0.000	22:47
Subject Sample # 1	0.178	22:48
Breath Volume	2.484	22:48
Air Blank	0.000	22:48
Air Blank	0.000	22:50
Subject Sample # 2	0.187	22:51
Breath Volume	3.117	22:51
Air Blank	0.000	22:52
Control Test	0.081	22:52
Air Blank	0.000	22:52
Diagnostics Check	OK	22:53

Data Download Date/Time: 05/11/2019 23:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: REIDY,T
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 04/20/2019 02:53:15

Diagnostics Check	OK	02:41	
Air Blank	0.000	02:42	
Control Test	0.082	02:42	
Air Blank	0.000	02:42	
Subject Sample	# 1	02:46	No Sample Provided
Breath Volume	0.000	02:46	
Air Blank	0.000	02:46	
Air Blank	0.000	02:48	
Subject Sample	# 2	02:51	Volume Not Met
Breath Volume	0.058	02:51	
Air Blank	0.000	02:52	
Control Test	0.082	02:52	
Air Blank	0.000	02:53	
Diagnostics Check	OK	02:53	

Data Download Date/Time: 05/11/2019 23:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CASTELEIRO,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: KOCH

DATE/TIME OF TEST: 04/20/2019 04:08:50

Diagnostics Check	OK	04:06	
Air Blank	0.000	04:06	
Control Test	0.082	04:06	
Air Blank	0.000	04:07	
Subject Sample	# 1	04:07	Subject Test Refused
Breath Volume	0.000	04:07	
Air Blank	0.000	04:07	
Control Test	0.082	04:08	
Air Blank	0.000	04:08	
Diagnostics Check	OK	04:08	

Data Download Date/Time: 05/11/2019 23:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,L
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: MCFEE

DATE/TIME OF TEST: 04/20/2019 04:13:47

Diagnostics Check	OK	04:11	
Air Blank	0.000	04:11	
Control Test	0.081	04:11	
Air Blank	0.000	04:12	
Subject Sample	# 1	04:12	Subject Test Refused
Breath Volume	0.000	04:12	
Air Blank	0.000	04:12	
Control Test	0.081	04:13	
Air Blank	0.000	04:13	
Diagnostics Check	OK	04:13	

Data Download Date/Time: 05/11/2019 23:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: WOODGET

DATE/TIME OF TEST: 04/21/2019 01:04:30

Diagnostics Check	OK	00:53	
Air Blank	0.000	00:53	
Control Test	0.082	00:53	
Air Blank	0.000	00:54	
Subject Sample	# 1	00:57	No Sample Provided
Breath Volume	0.000	00:57	
Air Blank	0.000	00:57	
Air Blank	0.000	00:59	
Subject Sample	# 2	01:03	No Sample Provided
Breath Volume	0.000	01:03	
Air Blank	0.000	01:03	
Control Test	0.082	01:03	
Air Blank	0.000	01:04	
Diagnostics Check	OK	01:04	

Data Download Date/Time: 05/11/2019 23:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEDE,BRENNEN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WEDE,BRENNEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: SKORCZEWSKI

DATE/TIME OF TEST: 04/21/2019 02:14:05

Diagnostics Check	OK	02:02	
Air Blank	0.000	02:02	
Control Test	0.081	02:03	
Air Blank	0.000	02:03	
Subject Sample	# 1 0.153	02:06	Volume Not Met
Breath Volume	0.050	02:06	
Air Blank	0.000	02:07	
Air Blank	0.000	02:09	
Subject Sample	# 2 0.146	02:12	Volume Not Met
Breath Volume	0.082	02:12	
Air Blank	0.000	02:13	
Control Test	0.080	02:13	
Air Blank	0.000	02:13	
Diagnostics Check	OK	02:14	

Data Download Date/Time: 05/11/2019 23:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEDE,BRENNEN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WEDE,BRENNEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: SKORCZEWSKI

DATE/TIME OF TEST: 04/21/2019 02:28:20

Diagnostics Check	OK	02:16	
Air Blank	0.000	02:17	
Control Test	0.080	02:17	
Air Blank	0.000	02:17	
Subject Sample	# 1 0.156	02:21	Volume Not Met
Breath Volume	0.062	02:21	
Air Blank	0.000	02:21	
Air Blank	0.000	02:23	
Subject Sample	# 2 0.159	02:26	Volume Not Met
Breath Volume	0.015	02:26	
Air Blank	0.000	02:27	
Control Test	0.078	02:27	
Air Blank	0.000	02:28	
Diagnostics Check	OK	02:28	

Data Download Date/Time: 05/11/2019 23:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEDE,BRENNEN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WEDE,BRENNEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: SKORCZEWSKI

DATE/TIME OF TEST: 04/21/2019 02:32:43

Diagnostics Check	OK	02:29	
Air Blank	0.000	02:30	
Control Test	0.080	02:30	
Air Blank	0.000	02:31	
Subject Sample	# 1	02:31	Subject Test Refused
Breath Volume	0.000	02:31	
Air Blank	0.000	02:31	
Control Test	0.080	02:32	
Air Blank	0.000	02:32	
Diagnostics Check	OK	02:32	

Data Download Date/Time: 05/11/2019 23:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: MALACARA VAZQUEZ

DATE/TIME OF TEST: 04/21/2019 04:10:28

Diagnostics Check	OK	04:03
Air Blank	0.000	04:03
Control Test	0.081	04:03
Air Blank	0.000	04:04
Subject Sample # 1	0.136	04:05
Breath Volume	2.660	04:05
Air Blank	0.000	04:06
Air Blank	0.000	04:08
Subject Sample # 2	0.134	04:08
Breath Volume	2.656	04:08
Air Blank	0.000	04:09
Control Test	0.081	04:09
Air Blank	0.000	04:10
Diagnostics Check	OK	04:10

Data Download Date/Time: 05/11/2019 23:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: BOTSFORD

DATE/TIME OF TEST: 04/22/2019 02:51:32

Diagnostics Check	OK	02:38	
Air Blank	0.000	02:39	
Control Test	0.082	02:39	
Air Blank	0.000	02:39	
Subject Sample	# 1	02:43	No Sample Provided
Breath Volume	0.000	02:43	
Air Blank	0.000	02:43	
Air Blank	0.000	02:45	
Subject Sample	# 2	0.228 02:46	
Breath Volume	2.863	02:46	
Air Blank	0.000	02:47	
Air Blank	0.000	02:49	
Subject Sample	# 3	0.214 02:50	
Breath Volume	2.140	02:50	
Air Blank	0.000	02:50	
Control Test	0.081	02:50	
Air Blank	0.000	02:51	
Diagnostics Check	OK	02:51	

Data Download Date/Time: 05/11/2019 23:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BYRD,WILLIAM
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: HEINE

DATE/TIME OF TEST: 04/22/2019 22:56:49

Diagnostics Check	OK	22:50
Air Blank	0.000	22:51
Control Test	0.081	22:51
Air Blank	0.000	22:52
Subject Sample # 1	0.274	22:52
Breath Volume	2.039	22:52
Air Blank	0.000	22:53
Air Blank	0.000	22:54
Subject Sample # 2	0.269	22:55
Breath Volume	1.480	22:55
Air Blank	0.000	22:55
Control Test	0.081	22:56
Air Blank	0.000	22:56
Diagnostics Check	OK	22:56

Data Download Date/Time: 05/11/2019 23:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAMES,STEVEN,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: JAMES,STEVEN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: AMARDEEP

DATE/TIME OF TEST: 04/23/2019 23:28:57

Diagnostics Check	OK	23:24	
Air Blank	0.000	23:24	
Control Test	0.082	23:25	
Air Blank	0.000	23:25	
Subject Sample	# 1	23:27	Subject Test Refused
Breath Volume	0.000	23:27	
Air Blank	0.000	23:28	
Control Test	0.082	23:28	
Air Blank	0.000	23:28	
Diagnostics Check	OK	23:28	

Data Download Date/Time: 05/11/2019 23:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CASTELEIRO,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: ELIZALDE

DATE/TIME OF TEST: 04/24/2019 22:31:17

Diagnostics Check	OK	22:23
Air Blank	0.000	22:24
Control Test	0.082	22:24
Air Blank	0.000	22:25
Subject Sample # 1	0.109	22:26
Breath Volume	1.972	22:26
Air Blank	0.000	22:26
Air Blank	0.000	22:28
Subject Sample # 2	0.108	22:29
Breath Volume	1.816	22:29
Air Blank	0.000	22:30
Control Test	0.081	22:30
Air Blank	0.000	22:31
Diagnostics Check	OK	22:31

Data Download Date/Time: 05/11/2019 23:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: YEARGIN,G
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 04/25/2019 03:37:43

Diagnostics Check	OK	03:31
Air Blank	0.000	03:31
Control Test	0.083	03:31
Air Blank	0.000	03:32
Subject Sample # 1	0.195	03:33
Breath Volume	1.328	03:33
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.201	03:36
Breath Volume	2.742	03:36
Air Blank	0.000	03:36
Control Test	0.081	03:37
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 05/11/2019 23:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: MARIANI

DATE/TIME OF TEST: 04/26/2019 00:31:15

Diagnostics Check	OK	00:28	
Air Blank	0.000	00:28	
Control Test	0.082	00:29	
Air Blank	0.000	00:29	
Subject Sample	# 1	00:29	Subject Test Refused
Breath Volume	0.000	00:29	
Air Blank	0.000	00:30	
Control Test	0.082	00:30	
Air Blank	0.000	00:31	
Diagnostics Check	OK	00:31	

Data Download Date/Time: 05/11/2019 23:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:58

SUBJECT'S LAST NAME: BOYETTE

DATE/TIME OF TEST: 04/26/2019 20:25:39

Diagnostics Check	OK	20:19
Air Blank	0.000	20:19
Control Test	0.082	20:19
Air Blank	0.000	20:20
Subject Sample # 1	0.044	20:21
Breath Volume	1.871	20:21
Air Blank	0.000	20:21
Air Blank	0.000	20:23
Subject Sample # 2	0.045	20:24
Breath Volume	2.472	20:24
Air Blank	0.000	20:24
Control Test	0.081	20:25
Air Blank	0.000	20:25
Diagnostics Check	OK	20:25

Data Download Date/Time: 05/11/2019 23:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: CALDERON

DATE/TIME OF TEST: 04/27/2019 00:04:29

Diagnostics Check	OK	00:01	
Air Blank	0.000	00:02	
Control Test	0.082	00:02	
Air Blank	0.000	00:02	
Subject Sample	# 1	00:03	Subject Test Refused
Breath Volume	0.000	00:03	
Air Blank	0.000	00:03	
Control Test	0.082	00:03	
Air Blank	0.000	00:04	
Diagnostics Check	OK	00:04	

Data Download Date/Time: 05/11/2019 23:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: REID

DATE/TIME OF TEST: 04/27/2019 03:37:52

Diagnostics Check	OK	03:31
Air Blank	0.000	03:32
Control Test	0.082	03:32
Air Blank	0.000	03:32
Subject Sample # 1	0.203	03:33
Breath Volume	1.679	03:33
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.189	03:36
Breath Volume	1.238	03:36
Air Blank	0.000	03:36
Control Test	0.082	03:37
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 05/11/2019 23:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 04/27/2019 04:38:26

Diagnostics Check	OK	04:35	
Air Blank	0.000	04:35	
Control Test	0.081	04:35	
Air Blank	0.000	04:36	
Subject Sample	# 1	04:37	Subject Test Refused
Breath Volume	0.000	04:37	
Air Blank	0.000	04:37	
Control Test	0.081	04:37	
Air Blank	0.000	04:38	
Diagnostics Check	OK	04:38	

Data Download Date/Time: 05/11/2019 23:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BYRD,W
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: SLATOR

DATE/TIME OF TEST: 04/27/2019 22:31:15

Diagnostics Check	OK	22:28	
Air Blank	0.000	22:28	
Control Test	0.082	22:28	
Air Blank	0.000	22:29	
Subject Sample	# 1	22:29	Subject Test Refused
Breath Volume	0.000	22:29	
Air Blank	0.000	22:30	
Control Test	0.082	22:30	
Air Blank	0.000	22:31	
Diagnostics Check	OK	22:31	

Data Download Date/Time: 05/11/2019 23:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVAREZ, YOEL-ALEXIZ,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ALVAREZ, YOEL-ALEXIZ
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: PARRISH

DATE/TIME OF TEST: 04/28/2019 00:13:50

Diagnostics Check	OK	00:07
Air Blank	0.000	00:08
Control Test	0.082	00:08
Air Blank	0.000	00:09
Subject Sample # 1	0.107	00:09
Breath Volume	2.621	00:09
Air Blank	0.000	00:10
Air Blank	0.000	00:11
Subject Sample # 2	0.107	00:12
Breath Volume	2.339	00:12
Air Blank	0.000	00:12
Control Test	0.082	00:13
Air Blank	0.000	00:13
Diagnostics Check	OK	00:13

Data Download Date/Time: 05/11/2019 23:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEDE,BRENNEN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WEDE,BRENNEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: GRAY

DATE/TIME OF TEST: 04/28/2019 03:26:11

Diagnostics Check	OK	03:19
Air Blank	0.000	03:19
Control Test	0.082	03:20
Air Blank	0.000	03:20
Subject Sample # 1	0.115	03:21
Breath Volume	1.746	03:21
Air Blank	0.000	03:22
Air Blank	0.000	03:23
Subject Sample # 2	0.113	03:24
Breath Volume	1.765	03:24
Air Blank	0.000	03:25
Control Test	0.082	03:25
Air Blank	0.000	03:26
Diagnostics Check	OK	03:26

Data Download Date/Time: 05/11/2019 23:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: NEYMAN

DATE/TIME OF TEST: 04/28/2019 04:04:38

Diagnostics Check	OK	03:57
Air Blank	0.000	03:57
Control Test	0.081	03:58
Air Blank	0.000	03:58
Subject Sample # 1	0.183	04:00
Breath Volume	1.308	04:00
Air Blank	0.000	04:00
Air Blank	0.000	04:02
Subject Sample # 2	0.176	04:03
Breath Volume	1.316	04:03
Air Blank	0.000	04:03
Control Test	0.081	04:04
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

Data Download Date/Time: 05/11/2019 23:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BELCHER,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:53

SUBJECT'S LAST NAME: LEWIS JR

DATE/TIME OF TEST: 04/28/2019 18:19:37

Diagnostics Check	OK	18:13
Air Blank	0.000	18:14
Control Test	0.078	18:14
Air Blank	0.000	18:14
Subject Sample # 1	0.237	18:15
Breath Volume	2.437	18:15
Air Blank	0.000	18:15
Air Blank	0.000	18:17
Subject Sample # 2	0.237	18:18
Breath Volume	1.750	18:18
Air Blank	0.000	18:18
Control Test	0.081	18:19
Air Blank	0.000	18:19
Diagnostics Check	OK	18:19

Data Download Date/Time: 05/11/2019 23:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,NELSON,X
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: SUTTON

DATE/TIME OF TEST: 04/30/2019 01:27:06

Diagnostics Check	OK	01:19
Air Blank	0.000	01:20
Control Test	0.083	01:20
Air Blank	0.000	01:20
Subject Sample # 1	0.087	01:22
Breath Volume	2.031	01:22
Air Blank	0.000	01:23
Air Blank	0.000	01:25
Subject Sample # 2	0.083	01:25
Breath Volume	1.496	01:25
Air Blank	0.000	01:26
Control Test	0.081	01:26
Air Blank	0.000	01:27
Diagnostics Check	OK	01:27

Data Download Date/Time: 05/11/2019 23:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KELLIHER,MICHAEL
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: MADRID

DATE/TIME OF TEST: 04/01/2019 02:35:50

Diagnostics Check	OK	02:24	
Air Blank	0.000	02:24	
Control Test	0.079	02:25	
Air Blank	0.000	02:25	
Subject Sample	# 1	02:28	No Sample Provided
Breath Volume	0.000	02:28	
Air Blank	0.000	02:29	
Air Blank	0.000	02:31	
Subject Sample	# 2	02:34	No Sample Provided
Breath Volume	0.000	02:34	
Air Blank	0.000	02:34	
Control Test	0.079	02:35	
Air Blank	0.000	02:35	
Diagnostics Check	OK	02:35	

Data Download Date/Time: 04/18/2019 17:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, YVETTE, D
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FOWLER, S
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 03/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 11:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:15

SUBJECT'S LAST NAME: BUCHANAN

DATE/TIME OF TEST: 04/02/2019 12:58:19

Diagnostics Check	OK	12:50
Air Blank	0.000	12:50
Control Test	0.079	12:50
Air Blank	0.000	12:51
Subject Sample # 1	0.000	12:53
Breath Volume	1.917	12:53
Air Blank	0.000	12:53
Air Blank	0.000	12:55
Subject Sample # 2	0.000	12:56
Breath Volume	2.449	12:56
Air Blank	0.000	12:57
Control Test	0.079	12:57
Air Blank	0.000	12:58
Diagnostics Check	OK	12:58

Data Download Date/Time: 04/19/2019 12:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KELLIHER,MICHAEL
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: MARTINEZ-PEREZ

DATE/TIME OF TEST: 04/02/2019 23:21:52

Diagnostics Check	OK	23:12
Air Blank	0.000	23:13
Control Test	0.079	23:13
Air Blank	0.000	23:14
Subject Sample # 1	0.225	23:16
Breath Volume	1.722	23:16
Air Blank	0.000	23:17
Air Blank	0.000	23:19
Subject Sample # 2	0.227	23:20
Breath Volume	2.871	23:20
Air Blank	0.000	23:20
Control Test	0.077	23:21
Air Blank	0.000	23:21
Diagnostics Check	OK	23:21

Data Download Date/Time: 04/19/2019 12:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LUNA,KRISTINA
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 03/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:28

SUBJECT'S LAST NAME: ESTRELLA MORALES

DATE/TIME OF TEST: 04/03/2019 21:55:32

Diagnostics Check	OK	21:49
Air Blank	0.000	21:49
Control Test	0.080	21:49
Air Blank	0.000	21:50
Subject Sample # 1	0.117	21:51
Breath Volume	1.906	21:51
Air Blank	0.000	21:51
Air Blank	0.000	21:53
Subject Sample # 2	0.110	21:54
Breath Volume	1.343	21:54
Air Blank	0.000	21:54
Control Test	0.079	21:54
Air Blank	0.000	21:55
Diagnostics Check	OK	21:55

Data Download Date/Time: 04/19/2019 12:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HOLLOWAY,DEVARIOUS
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 03/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: COLAFERRI

DATE/TIME OF TEST: 04/05/2019 03:33:00

Diagnostics Check	OK	03:22
Air Blank	0.000	03:23
Control Test	0.079	03:23
Air Blank	0.000	03:24
Subject Sample # 1	0.189	03:25
Breath Volume	1.171	03:25
Air Blank	0.000	03:25
Air Blank	0.000	03:27
Subject Sample # 2	0.218	03:28
Breath Volume	2.941	03:28
Air Blank	0.000	03:29
Air Blank	0.000	03:30
Subject Sample # 3	0.213	03:31
Breath Volume	2.875	03:31
Air Blank	0.000	03:32
Control Test	0.077	03:32
Air Blank	0.000	03:32
Diagnostics Check	OK	03:33

Data Download Date/Time: 04/19/2019 12:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 18:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: WORKMAN

DATE/TIME OF TEST: 04/06/2019 19:42:12

Diagnostics Check	OK	19:35
Air Blank	0.000	19:36
Control Test	0.079	19:36
Air Blank	0.000	19:36
Subject Sample # 1	0.312	19:37
Breath Volume	1.246	19:37
Air Blank	0.000	19:38
Air Blank	0.000	19:40
Subject Sample # 2	0.306	19:40
Breath Volume	1.273	19:40
Air Blank	0.000	19:41
Control Test	0.079	19:41
Air Blank	0.000	19:42
Diagnostics Check	OK	19:42

Data Download Date/Time: 04/19/2019 12:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BLANDING,ROBERT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 04/06/2019 23:58:24

Diagnostics Check	OK	23:52
Air Blank	0.000	23:52
Control Test	0.079	23:52
Air Blank	0.000	23:53
Subject Sample # 1	0.176	23:53
Breath Volume	1.906	23:53
Air Blank	0.000	23:54
Air Blank	0.000	23:56
Subject Sample # 2	0.173	23:56
Breath Volume	1.445	23:56
Air Blank	0.000	23:57
Control Test	0.079	23:57
Air Blank	0.000	23:58
Diagnostics Check	OK	23:58

Data Download Date/Time: 04/19/2019 12:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HOLLOWAY,DEVARIOUS
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 03/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: MONDESIR

DATE/TIME OF TEST: 04/07/2019 04:28:55

Diagnostics Check	OK	04:22
Air Blank	0.000	04:23
Control Test	0.079	04:23
Air Blank	0.000	04:23
Subject Sample # 1	0.219	04:24
Breath Volume	1.886	04:24
Air Blank	0.000	04:24
Air Blank	0.000	04:26
Subject Sample # 2	0.205	04:27
Breath Volume	2.058	04:27
Air Blank	0.000	04:28
Control Test	0.079	04:28
Air Blank	0.000	04:28
Diagnostics Check	OK	04:28

Data Download Date/Time: 04/19/2019 12:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SPROUSE,MICHAEL
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 03/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 05:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:17

SUBJECT'S LAST NAME: PEREIRA DOS SANTOS

DATE/TIME OF TEST: 04/07/2019 06:56:02

Diagnostics Check	OK	06:49
Air Blank	0.000	06:50
Control Test	0.079	06:50
Air Blank	0.000	06:51
Subject Sample # 1	0.018	06:51
Breath Volume	2.355	06:51
Air Blank	0.000	06:52
Air Blank	0.000	06:54
Subject Sample # 2	0.016	06:54
Breath Volume	1.203	06:54
Air Blank	0.000	06:55
Control Test	0.079	06:55
Air Blank	0.000	06:55
Diagnostics Check	OK	06:56

Data Download Date/Time: 04/19/2019 12:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BERGER,MICHAEL
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 03/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 04/11/2019 01:56:46

Diagnostics Check	OK	01:50
Air Blank	0.000	01:50
Control Test	0.079	01:50
Air Blank	0.000	01:51
Subject Sample # 1	0.286	01:52
Breath Volume	1.871	01:52
Air Blank	0.000	01:52
Air Blank	0.000	01:54
Subject Sample # 2	0.279	01:55
Breath Volume	1.675	01:55
Air Blank	0.000	01:55
Control Test	0.078	01:56
Air Blank	0.000	01:56
Diagnostics Check	OK	01:56

Data Download Date/Time: 04/19/2019 12:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FITHIAN,JASON
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 03/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: BANKS

DATE/TIME OF TEST: 04/13/2019 02:16:17

Diagnostics Check	OK	02:07
Air Blank	0.000	02:08
Control Test	0.079	02:08
Air Blank	0.000	02:09
Subject Sample # 1	0.168	02:10
Breath Volume	1.437	02:10
Air Blank	0.000	02:11
Air Blank	0.000	02:13
Subject Sample # 2	0.164	02:14
Breath Volume	1.621	02:14
Air Blank	0.000	02:15
Control Test	0.079	02:15
Air Blank	0.000	02:16
Diagnostics Check	OK	02:16

Data Download Date/Time: 04/19/2019 12:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FITHIAN,JASON
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 03/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 04:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 04/13/2019 05:07:36

Diagnostics Check	OK	05:00
Air Blank	0.000	05:01
Control Test	0.078	05:01
Air Blank	0.000	05:02
Subject Sample # 1	0.223	05:02
Breath Volume	2.546	05:02
Air Blank	0.000	05:03
Air Blank	0.000	05:05
Subject Sample # 2	0.203	05:06
Breath Volume	1.976	05:06
Air Blank	0.000	05:06
Control Test	0.078	05:07
Air Blank	0.000	05:07
Diagnostics Check	OK	05:07

Data Download Date/Time: 04/19/2019 12:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CLAIRE,CASEY
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 03/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 04:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:41

SUBJECT'S LAST NAME: CORTES-CORTES

DATE/TIME OF TEST: 04/13/2019 06:09:21

Diagnostics Check	OK	06:02
Air Blank	0.000	06:02
Control Test	0.079	06:03
Air Blank	0.000	06:03
Subject Sample # 1	0.128	06:04
Breath Volume	1.695	06:04
Air Blank	0.000	06:04
Air Blank	0.000	06:06
Subject Sample # 2	0.133	06:07
Breath Volume	1.234	06:07
Air Blank	0.000	06:08
Control Test	0.080	06:08
Air Blank	0.000	06:09
Diagnostics Check	OK	06:09

Data Download Date/Time: 04/19/2019 12:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: BOSSELMAN

DATE/TIME OF TEST: 04/14/2019 01:32:34

Diagnostics Check	OK	01:26
Air Blank	0.000	01:26
Control Test	0.079	01:27
Air Blank	0.000	01:27
Subject Sample # 1	0.188	01:28
Breath Volume	1.437	01:28
Air Blank	0.000	01:28
Air Blank	0.000	01:30
Subject Sample # 2	0.185	01:31
Breath Volume	1.152	01:31
Air Blank	0.000	01:31
Control Test	0.079	01:31
Air Blank	0.000	01:32
Diagnostics Check	OK	01:32

Data Download Date/Time: 04/19/2019 12:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HOLLOWAY,DEVARIOUS
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 03/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: AMBROISE

DATE/TIME OF TEST: 04/14/2019 04:11:06

Diagnostics Check	OK	04:05
Air Blank	0.000	04:05
Control Test	0.079	04:05
Air Blank	0.000	04:06
Subject Sample # 1	0.109	04:06
Breath Volume	2.062	04:06
Air Blank	0.000	04:07
Air Blank	0.000	04:09
Subject Sample # 2	0.109	04:09
Breath Volume	2.828	04:09
Air Blank	0.000	04:10
Control Test	0.079	04:10
Air Blank	0.000	04:11
Diagnostics Check	OK	04:11

Data Download Date/Time: 04/19/2019 12:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: STONER,NATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 03/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: OLIVO ROSARIO

DATE/TIME OF TEST: 04/15/2019 01:42:42

Diagnostics Check	OK	01:36
Air Blank	0.000	01:37
Control Test	0.078	01:37
Air Blank	0.000	01:37
Subject Sample # 1	0.155	01:38
Breath Volume	2.515	01:38
Air Blank	0.000	01:38
Air Blank	0.000	01:40
Subject Sample # 2	0.154	01:41
Breath Volume	1.882	01:41
Air Blank	0.000	01:41
Control Test	0.079	01:42
Air Blank	0.000	01:42
Diagnostics Check	OK	01:42

Data Download Date/Time: 04/19/2019 12:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ORTUNO,CARLOS
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 03/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: PAREDES MEJIA

DATE/TIME OF TEST: 04/15/2019 02:45:32

Diagnostics Check	OK	02:39
Air Blank	0.000	02:39
Control Test	0.079	02:40
Air Blank	0.000	02:40
Subject Sample # 1	0.144	02:41
Breath Volume	2.496	02:41
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.135	02:44
Breath Volume	1.226	02:44
Air Blank	0.000	02:44
Control Test	0.079	02:44
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 04/19/2019 12:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: AKERS,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 08:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:59

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 04/15/2019 10:30:13

Diagnostics Check	OK	10:24
Air Blank	0.000	10:24
Control Test	0.079	10:24
Air Blank	0.000	10:25
Subject Sample # 1	0.000	10:25
Breath Volume	1.843	10:25
Air Blank	0.000	10:26
Air Blank	0.000	10:28
Subject Sample # 2	0.000	10:28
Breath Volume	1.613	10:28
Air Blank	0.000	10:29
Control Test	0.080	10:29
Air Blank	0.000	10:30
Diagnostics Check	OK	10:30

Data Download Date/Time: 04/19/2019 12:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KELLIHER,MICHAEL
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 04/18/2019 04:09:12

Diagnostics Check	OK	04:06	
Air Blank	0.000	04:06	
Control Test	0.079	04:07	
Air Blank	0.000	04:07	
Subject Sample	# 1	04:07	Subject Test Refused
Breath Volume	0.000	04:07	
Air Blank	0.000	04:08	
Control Test	0.079	04:08	
Air Blank	0.000	04:09	
Diagnostics Check	OK	04:09	

Data Download Date/Time: 04/19/2019 12:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: NISTER,JOSPEH
ARRESTING OFFICER AGENCY: Lauderhill PD
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: DUARTE

DATE/TIME OF TEST: 04/19/2019 04:03:38

Diagnostics Check	OK	03:56
Air Blank	0.000	03:56
Control Test	0.080	03:56
Air Blank	0.000	03:57
Subject Sample # 1	0.203	03:58
Breath Volume	1.886	03:58
Air Blank	0.000	03:58
Air Blank	0.000	04:00
Subject Sample # 2	0.185	04:02
Breath Volume	1.769	04:02
Air Blank	0.000	04:02
Control Test	0.079	04:03
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time: 04/19/2019 12:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MASTERSON,MONIQUE
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 04:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:11

SUBJECT'S LAST NAME: VELASQUEZ MERNDEZ

DATE/TIME OF TEST: 04/19/2019 05:50:06

Diagnostics Check	OK	05:36	
Air Blank	0.000	05:36	
Control Test	0.080	05:36	
Air Blank	0.000	05:37	
Subject Sample	# 1 0.191	05:40	Volume Not Met
Breath Volume	0.089	05:40	
Air Blank	0.000	05:41	
Air Blank	0.000	05:42	
Subject Sample	# 2 0.196	05:43	
Breath Volume	1.105	05:43	
Air Blank	0.000	05:44	
Air Blank	0.000	05:46	
Subject Sample	# 3 0.190	05:48	
Breath Volume	2.136	05:48	
Air Blank	0.000	05:49	
Control Test	0.078	05:49	
Air Blank	0.000	05:50	
Diagnostics Check	OK	05:50	

Data Download Date/Time: 04/19/2019 12:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: IRELAN,ANDREW
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: CUNNINGHAM

DATE/TIME OF TEST: 04/20/2019 03:17:48

Diagnostics Check	OK	03:10
Air Blank	0.000	03:11
Control Test	0.080	03:11
Air Blank	0.000	03:12
Subject Sample # 1	0.188	03:12
Breath Volume	1.410	03:12
Air Blank	0.000	03:13
Air Blank	0.000	03:15
Subject Sample # 2	0.178	03:16
Breath Volume	1.804	03:16
Air Blank	0.000	03:16
Control Test	0.079	03:17
Air Blank	0.000	03:17
Diagnostics Check	OK	03:17

Data Download Date/Time: 05/09/2019 13:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KEATING,KAYLA
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 04:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: BEESON

DATE/TIME OF TEST: 04/20/2019 05:05:41

Diagnostics Check	OK	04:59
Air Blank	0.000	05:00
Control Test	0.079	05:00
Air Blank	0.000	05:00
Subject Sample # 1	0.000	05:01
Breath Volume	2.605	05:01
Air Blank	0.000	05:01
Air Blank	0.000	05:03
Subject Sample # 2	0.000	05:04
Breath Volume	2.476	05:04
Air Blank	0.000	05:04
Control Test	0.079	05:05
Air Blank	0.000	05:05
Diagnostics Check	OK	05:05

Data Download Date/Time: 05/09/2019 13:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: STONER,NATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: GALLAHER

DATE/TIME OF TEST: 04/21/2019 01:20:54

Diagnostics Check	OK	01:14
Air Blank	0.000	01:14
Control Test	0.079	01:14
Air Blank	0.000	01:15
Subject Sample # 1	0.096	01:16
Breath Volume	1.464	01:16
Air Blank	0.000	01:16
Air Blank	0.000	01:18
Subject Sample # 2	0.093	01:19
Breath Volume	1.472	01:19
Air Blank	0.000	01:19
Control Test	0.079	01:20
Air Blank	0.000	01:20
Diagnostics Check	OK	01:20

Data Download Date/Time: 05/09/2019 13:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: IRELAN,ANDREW
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: NGO

DATE/TIME OF TEST: 04/21/2019 05:11:38

Diagnostics Check	OK	05:04
Air Blank	0.000	05:05
Control Test	0.080	05:05
Air Blank	0.000	05:06
Subject Sample # 1	0.055	05:06
Breath Volume	1.164	05:06
Air Blank	0.000	05:07
Air Blank	0.000	05:09
Subject Sample # 2	0.052	05:10
Breath Volume	1.628	05:10
Air Blank	0.000	05:10
Control Test	0.079	05:11
Air Blank	0.000	05:11
Diagnostics Check	OK	05:11

Data Download Date/Time: 05/09/2019 13:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: STONER,NATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 04/21/2019 23:55:28

Diagnostics Check	OK	23:49
Air Blank	0.000	23:49
Control Test	0.079	23:49
Air Blank	0.000	23:50
Subject Sample # 1	0.191	23:50
Breath Volume	1.531	23:50
Air Blank	0.000	23:51
Air Blank	0.000	23:53
Subject Sample # 2	0.178	23:53
Breath Volume	1.601	23:53
Air Blank	0.000	23:54
Control Test	0.079	23:54
Air Blank	0.000	23:55
Diagnostics Check	OK	23:55

Data Download Date/Time: 05/09/2019 13:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HOLLOWAY,DEVARIOUS
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: PALMER

DATE/TIME OF TEST: 04/24/2019 01:34:36

Diagnostics Check	OK	01:25
Air Blank	0.000	01:26
Control Test	0.079	01:26
Air Blank	0.000	01:26
Subject Sample # 1	0.150	01:29
Breath Volume	1.593	01:29
Air Blank	0.000	01:30
Air Blank	0.000	01:32
Subject Sample # 2	0.152	01:33
Breath Volume	1.257	01:33
Air Blank	0.000	01:33
Control Test	0.079	01:34
Air Blank	0.000	01:34
Diagnostics Check	OK	01:34

Data Download Date/Time: 05/09/2019 13:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FITHIAN,JASON
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 04/26/2019 02:12:49

Diagnostics Check	OK	02:06
Air Blank	0.000	02:06
Control Test	0.079	02:07
Air Blank	0.000	02:07
Subject Sample # 1	0.138	02:08
Breath Volume	1.835	02:08
Air Blank	0.000	02:08
Air Blank	0.000	02:10
Subject Sample # 2	0.130	02:11
Breath Volume	2.402	02:11
Air Blank	0.000	02:11
Control Test	0.078	02:12
Air Blank	0.000	02:12
Diagnostics Check	OK	02:12

Data Download Date/Time: 05/09/2019 13:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ROJAS,RANDY
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 04:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:01

SUBJECT'S LAST NAME: UVA

DATE/TIME OF TEST: 04/26/2019 05:30:38

Diagnostics Check	OK	05:23
Air Blank	0.000	05:24
Control Test	0.079	05:24
Air Blank	0.000	05:25
Subject Sample # 1	0.158	05:25
Breath Volume	2.847	05:25
Air Blank	0.000	05:26
Air Blank	0.000	05:28
Subject Sample # 2	0.163	05:29
Breath Volume	1.187	05:29
Air Blank	0.000	05:29
Control Test	0.079	05:30
Air Blank	0.000	05:30
Diagnostics Check	OK	05:30

Data Download Date/Time: 05/09/2019 13:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FRANK,TAYLOR
ARRESTING OFFICER AGENCY: Seminole Dept of LE
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: GOPHER JR

DATE/TIME OF TEST: 04/29/2019 00:12:48

Diagnostics Check	OK	00:06
Air Blank	0.000	00:07
Control Test	0.079	00:07
Air Blank	0.000	00:07
Subject Sample # 1	0.278	00:08
Breath Volume	2.742	00:08
Air Blank	0.000	00:08
Air Blank	0.000	00:10
Subject Sample # 2	0.265	00:11
Breath Volume	2.750	00:11
Air Blank	0.000	00:11
Control Test	0.078	00:12
Air Blank	0.000	00:12
Diagnostics Check	OK	00:12

Data Download Date/Time: 05/09/2019 13:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: STEELE,MICHAEL
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 07:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:20

SUBJECT'S LAST NAME: ANGILOT

DATE/TIME OF TEST: 04/29/2019 09:50:14

Diagnostics Check	OK	09:41
Air Blank	0.000	09:42
Control Test	0.080	09:42
Air Blank	0.000	09:43
Subject Sample # 1	0.227	09:44
Breath Volume	1.433	09:44
Air Blank	0.000	09:44
Air Blank	0.000	09:46
Subject Sample # 2	0.223	09:48
Breath Volume	1.148	09:48
Air Blank	0.000	09:49
Control Test	0.079	09:49
Air Blank	0.000	09:50
Diagnostics Check	OK	09:50

Data Download Date/Time: 05/09/2019 13:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARING,BRANDON,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: RODRIGUEZ,XAVIER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 03/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: MONROY-RESENDIZ

DATE/TIME OF TEST: 04/01/2019 22:01:45

Diagnostics Check	OK	21:57	
Air Blank	0.000	21:58	
Control Test	0.079	21:58	
Air Blank	0.000	21:59	
Subject Sample	# 1 0.221	22:00	
Breath Volume	1.699	22:00	
Air Blank		22:01	Purge Fail
Air Blank	0.000	22:01	

Data Download Date/Time: 05/02/2019 17:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARING,BRANDON,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: RODRIGUEZ,XAVIER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 03/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: MONROY-RESENDIZ

DATE/TIME OF TEST: 04/01/2019 22:11:18

Diagnostics Check	OK	22:03
Air Blank	0.000	22:04
Control Test	0.076	22:04
Air Blank	0.000	22:05
Subject Sample # 1	0.213	22:06
Breath Volume	1.273	22:06
Air Blank	0.000	22:07
Air Blank	0.000	22:08
Subject Sample # 2	0.196	22:09
Breath Volume	1.570	22:09
Air Blank	0.000	22:10
Control Test	0.076	22:10
Air Blank	0.000	22:11
Diagnostics Check	OK	22:11

Data Download Date/Time: 05/02/2019 17:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDMONDS,RICHARD,J
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: MEDLIN,NAT
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 03/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: BONHAM

DATE/TIME OF TEST: 04/04/2019 01:28:41

Diagnostics Check	OK	01:22
Air Blank	0.000	01:22
Control Test	0.079	01:23
Air Blank	0.000	01:23
Subject Sample # 1	0.142	01:24
Breath Volume	3.378	01:24
Air Blank	0.000	01:24
Air Blank	0.000	01:26
Subject Sample # 2	0.145	01:26
Breath Volume	1.675	01:26
Air Blank	0.000	01:27
Control Test	0.078	01:28
Air Blank	0.000	01:28
Diagnostics Check	OK	01:28

Data Download Date/Time: 05/02/2019 17:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEGERSMITH,DANIEL,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,DANIEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 03/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: MC KAMEY

DATE/TIME OF TEST: 04/05/2019 01:10:04

Diagnostics Check	OK	00:54	
Air Blank	0.000	00:55	
Control Test	0.079	00:55	
Air Blank	0.000	00:56	
Subject Sample	# 1 0.199	00:59	Volume Not Met
Breath Volume	0.269	00:59	
Air Blank	0.000	01:00	
Air Blank	0.000	01:01	
Subject Sample	# 2 0.184	01:02	No .020 agreement
Breath Volume	1.570	01:02	
Air Blank	0.000	01:03	
Air Blank	0.000	01:05	
Subject Sample	# 3 0.205	01:08	No .020 agreement
Breath Volume	1.839	01:08	
Air Blank	0.000	01:09	
Control Test	0.076	01:09	
Air Blank	0.000	01:09	
Diagnostics Check	OK	01:10	

Data Download Date/Time: 05/02/2019 17:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HURT,JONATHAN,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: HURT,JONATHAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 03/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: RAY

DATE/TIME OF TEST: 04/05/2019 03:08:06

Diagnostics Check	OK	03:00
Air Blank	0.000	03:00
Control Test	0.079	03:01
Air Blank	0.000	03:01
Subject Sample # 1	0.154	03:02
Breath Volume	5.003	03:02
Air Blank	0.000	03:03
Air Blank	0.000	03:05
Subject Sample # 2	0.155	03:06
Breath Volume	4.984	03:06
Air Blank	0.000	03:07
Control Test	0.078	03:07
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

Data Download Date/Time: 05/02/2019 17:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUIS,J.,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: BUIS,J.
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 03/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: CARIDI

DATE/TIME OF TEST: 04/06/2019 01:17:59

Diagnostics Check	OK	01:10
Air Blank	0.000	01:11
Control Test	0.079	01:11
Air Blank	0.000	01:12
Subject Sample # 1	0.140	01:13
Breath Volume	2.121	01:13
Air Blank	0.000	01:13
Air Blank	0.000	01:15
Subject Sample # 2	0.129	01:16
Breath Volume	1.410	01:16
Air Blank	0.000	01:17
Control Test	0.079	01:17
Air Blank	0.000	01:17
Diagnostics Check	OK	01:17

Data Download Date/Time: 05/02/2019 17:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HURT,JONATHAN,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: HURT,JONATHAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 03/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: RUMAK

DATE/TIME OF TEST: 04/06/2019 01:51:25

Diagnostics Check	OK	01:44
Air Blank	0.000	01:44
Control Test	0.079	01:44
Air Blank	0.000	01:45
Subject Sample # 1	0.249	01:46
Breath Volume	1.402	01:46
Air Blank	0.000	01:47
Air Blank	0.000	01:48
Subject Sample # 2	0.267	01:49
Breath Volume	2.324	01:49
Air Blank	0.000	01:50
Control Test	0.079	01:50
Air Blank	0.000	01:51
Diagnostics Check	OK	01:51

Data Download Date/Time: 05/02/2019 17:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUGENT,DAVID,A
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NUGENT,DAVID
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 03/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:49

SUBJECT'S LAST NAME: GUTIERREZ

DATE/TIME OF TEST: 04/06/2019 18:18:57

Diagnostics Check	OK	18:11
Air Blank	0.000	18:11
Control Test	0.079	18:12
Air Blank	0.000	18:12
Subject Sample # 1	0.150	18:13
Breath Volume	1.617	18:13
Air Blank	0.000	18:14
Air Blank	0.000	18:15
Subject Sample # 2	0.155	18:17
Breath Volume	2.753	18:17
Air Blank	0.000	18:18
Control Test	0.079	18:18
Air Blank	0.000	18:18
Diagnostics Check	OK	18:18

Data Download Date/Time: 05/02/2019 17:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUGENT,DAVID,A
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NUGENT,DAVID
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 03/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 10:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:21

SUBJECT'S LAST NAME: FRENCH

DATE/TIME OF TEST: 04/10/2019 12:02:32

Diagnostics Check	OK	11:51	
Air Blank	0.000	11:51	
Control Test	0.079	11:51	
Air Blank	0.000	11:52	
Subject Sample	# 1	11:55	No Sample Provided
Breath Volume	0.000	11:55	
Air Blank	0.000	11:56	
Air Blank	0.000	11:58	
Subject Sample	# 2	12:01	No Sample Provided
Breath Volume	0.000	12:01	
Air Blank	0.000	12:01	
Control Test	0.079	12:01	
Air Blank	0.000	12:02	
Diagnostics Check	OK	12:02	

Data Download Date/Time: 05/02/2019 17:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HURT,JONATHAN,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: HURT,JONATHAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 03/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:17

SUBJECT'S LAST NAME: TRUITT

DATE/TIME OF TEST: 04/12/2019 19:25:26

Diagnostics Check	OK	19:13	
Air Blank	0.000	19:14	
Control Test	0.079	19:14	
Air Blank	0.000	19:15	
Subject Sample	# 1	19:18	No Sample Provided
Breath Volume	0.000	19:18	
Air Blank	0.000	19:18	
Air Blank	0.000	19:20	
Subject Sample	# 2	19:24	No Sample Provided
Breath Volume	0.000	19:24	
Air Blank	0.000	19:24	
Control Test	0.079	19:24	
Air Blank	0.000	19:25	
Diagnostics Check	OK	19:25	

Data Download Date/Time: 05/02/2019 17:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUIS,OFC. J,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: BUIS,OFC. J
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 03/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: VERT

DATE/TIME OF TEST: 04/13/2019 02:00:16

Diagnostics Check	OK	01:53
Air Blank	0.000	01:54
Control Test	0.079	01:54
Air Blank	0.000	01:55
Subject Sample # 1	0.087	01:55
Breath Volume	2.601	01:55
Air Blank	0.000	01:56
Air Blank	0.000	01:58
Subject Sample # 2	0.082	01:58
Breath Volume	1.710	01:58
Air Blank	0.000	01:59
Control Test	0.078	01:59
Air Blank	0.000	02:00
Diagnostics Check	OK	02:00

Data Download Date/Time: 05/02/2019 17:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNEAU,OFC,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: BEAVER,OFC
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 03/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:37

SUBJECT'S LAST NAME: WAITS

DATE/TIME OF TEST: 04/17/2019 21:15:49

Diagnostics Check	OK	21:09
Air Blank	0.000	21:09
Control Test	0.079	21:09
Air Blank	0.000	21:10
Subject Sample # 1	0.045	21:11
Breath Volume	4.296	21:11
Air Blank	0.000	21:11
Air Blank	0.000	21:13
Subject Sample # 2	0.043	21:14
Breath Volume	4.640	21:14
Air Blank	0.000	21:14
Control Test	0.079	21:15
Air Blank	0.000	21:15
Diagnostics Check	OK	21:15

Data Download Date/Time: 05/02/2019 17:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HURT,JONATHAN,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: CORUJO,EMMANUEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 03/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:06

SUBJECT'S LAST NAME: BRANDS

DATE/TIME OF TEST: 04/20/2019 19:01:25

Diagnostics Check	OK	18:54
Air Blank	0.000	18:54
Control Test	0.078	18:55
Air Blank	0.000	18:55
Subject Sample # 1	0.220	18:56
Breath Volume	2.710	18:56
Air Blank	0.000	18:57
Air Blank	0.000	18:58
Subject Sample # 2	0.213	18:59
Breath Volume	2.613	18:59
Air Blank	0.000	19:00
Control Test	0.078	19:00
Air Blank	0.000	19:01
Diagnostics Check	OK	19:01

Data Download Date/Time: 05/02/2019 17:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDMONDS,RICHARD,J
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: WILLIAMS,JEFF
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 03/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: AVALOS-ANGELES

DATE/TIME OF TEST: 04/22/2019 01:39:15

Diagnostics Check	OK	01:32
Air Blank	0.000	01:32
Control Test	0.079	01:33
Air Blank	0.000	01:33
Subject Sample # 1	0.212	01:34
Breath Volume	2.121	01:34
Air Blank	0.000	01:35
Air Blank	0.000	01:37
Subject Sample # 2	0.211	01:37
Breath Volume	1.832	01:37
Air Blank	0.000	01:38
Control Test	0.078	01:38
Air Blank	0.000	01:39
Diagnostics Check	OK	01:39

Data Download Date/Time: 05/02/2019 17:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDMONDS,RICHARD,J
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: YUEN,JOHN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 03/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: PERAMAS

DATE/TIME OF TEST: 04/22/2019 04:07:59

Diagnostics Check	OK	04:01
Air Blank	0.000	04:01
Control Test	0.079	04:01
Air Blank	0.000	04:02
Subject Sample # 1	0.148	04:03
Breath Volume	1.554	04:03
Air Blank	0.000	04:03
Air Blank	0.000	04:05
Subject Sample # 2	0.156	04:06
Breath Volume	1.941	04:06
Air Blank	0.000	04:07
Control Test	0.080	04:07
Air Blank	0.000	04:07
Diagnostics Check	OK	04:07

Data Download Date/Time: 05/02/2019 17:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THEODORE,J
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: MILLER,THEODORE
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 03/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: BACHA

DATE/TIME OF TEST: 04/23/2019 23:53:19

Diagnostics Check	OK	23:46
Air Blank	0.000	23:47
Control Test	0.080	23:47
Air Blank	0.000	23:47
Subject Sample # 1	0.131	23:48
Breath Volume	2.101	23:48
Air Blank	0.000	23:48
Air Blank	0.000	23:50
Subject Sample # 2	0.141	23:51
Breath Volume	3.519	23:51
Air Blank	0.000	23:52
Control Test	0.081	23:52
Air Blank	0.000	23:53
Diagnostics Check	OK	23:53

Data Download Date/Time: 05/02/2019 17:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,CHAD,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: REED,CHAD
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 03/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: PAGE

DATE/TIME OF TEST: 04/25/2019 21:23:53

Diagnostics Check	OK	21:16
Air Blank	0.000	21:17
Control Test	0.078	21:17
Air Blank	0.000	21:18
Subject Sample # 1	0.143	21:19
Breath Volume	1.875	21:19
Air Blank	0.000	21:19
Air Blank	0.000	21:21
Subject Sample # 2	0.144	21:22
Breath Volume	1.539	21:22
Air Blank	0.000	21:22
Control Test	0.079	21:23
Air Blank	0.000	21:23
Diagnostics Check	OK	21:23

Data Download Date/Time: 05/02/2019 17:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICHMOND, MATTHEW, D
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: DIAZ, M
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 03/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 08:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:32

SUBJECT'S LAST NAME: RILEY

DATE/TIME OF TEST: 04/26/2019 09:59:21

Diagnostics Check	OK	09:52
Air Blank	0.000	09:52
Control Test	0.079	09:53
Air Blank	0.000	09:53
Subject Sample # 1	0.017	09:54
Breath Volume	1.957	09:54
Air Blank	0.000	09:55
Air Blank	0.000	09:57
Subject Sample # 2	0.017	09:57
Breath Volume	2.226	09:57
Air Blank	0.000	09:58
Control Test	0.079	09:58
Air Blank	0.000	09:59
Diagnostics Check	OK	09:59

Data Download Date/Time: 05/02/2019 17:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OBRIEN,DAN,P
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: MILLS,ANTONY
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: ESPANOL

DATE/TIME OF TEST: 04/27/2019 04:14:42

Diagnostics Check	OK	04:07
Air Blank	0.000	04:08
Control Test	0.079	04:08
Air Blank	0.000	04:09
Subject Sample # 1	0.196	04:09
Breath Volume	2.945	04:09
Air Blank	0.000	04:10
Air Blank	0.000	04:12
Subject Sample # 2	0.196	04:12
Breath Volume	3.476	04:12
Air Blank	0.000	04:13
Control Test	0.079	04:14
Air Blank	0.000	04:14
Diagnostics Check	OK	04:14

Data Download Date/Time: 05/02/2019 17:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEGERSMITH,DANIEL,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,DANIEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:01

SUBJECT'S LAST NAME: CARIDE

DATE/TIME OF TEST: 04/27/2019 19:32:17

Diagnostics Check	OK	19:25
Air Blank	0.000	19:25
Control Test	0.079	19:26
Air Blank	0.000	19:26
Subject Sample # 1	0.070	19:27
Breath Volume	2.019	19:27
Air Blank	0.000	19:28
Air Blank	0.000	19:29
Subject Sample # 2	0.069	19:30
Breath Volume	1.996	19:30
Air Blank	0.000	19:31
Control Test	0.080	19:31
Air Blank	0.000	19:32
Diagnostics Check	OK	19:32

Data Download Date/Time: 05/02/2019 17:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUGENT,DAVID,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NUGENT,DAVID
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/26/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: KIRN

DATE/TIME OF TEST: 04/28/2019 00:59:34

Diagnostics Check	OK	00:53
Air Blank	0.000	00:53
Control Test	0.080	00:53
Air Blank	0.000	00:54
Subject Sample # 1	0.131	00:55
Breath Volume	4.253	00:55
Air Blank	0.000	00:55
Air Blank	0.000	00:57
Subject Sample # 2	0.126	00:57
Breath Volume	2.421	00:57
Air Blank	0.000	00:58
Control Test	0.079	00:58
Air Blank	0.000	00:59
Diagnostics Check	OK	00:59

Data Download Date/Time: 05/02/2019 17:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OBRIEN,DANIEL,P
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: MILLS,ANTONY
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:54

SUBJECT'S LAST NAME: FOSCO

DATE/TIME OF TEST: 04/28/2019 20:24:52

Diagnostics Check	OK	20:17
Air Blank	0.000	20:17
Control Test	0.078	20:17
Air Blank	0.000	20:18
Subject Sample # 1	0.253	20:19
Breath Volume	3.562	20:19
Air Blank	0.000	20:20
Air Blank	0.000	20:22
Subject Sample # 2	0.257	20:23
Breath Volume	2.746	20:23
Air Blank	0.000	20:23
Control Test	0.078	20:24
Air Blank	0.000	20:24
Diagnostics Check	OK	20:24

Data Download Date/Time: 05/02/2019 17:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,BARRY,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: NIXON,BARRY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: BORDA-VILA

DATE/TIME OF TEST: 04/12/2019 23:34:51

Diagnostics Check	OK	23:28
Air Blank	0.000	23:28
Control Test	0.077	23:29
Air Blank	0.000	23:29
Subject Sample # 1	0.144	23:29
Breath Volume	1.421	23:29
Air Blank	0.000	23:30
Air Blank	0.000	23:32
Subject Sample # 2	0.144	23:33
Breath Volume	3.214	23:33
Air Blank	0.000	23:33
Control Test	0.077	23:34
Air Blank	0.000	23:34
Diagnostics Check	OK	23:34

Data Download Date/Time: 04/27/2019 12:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH-LONG,JEFFERSON,D
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: BROWN,JEFFERY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: THURSTON

DATE/TIME OF TEST: 04/17/2019 23:16:56

Diagnostics Check	OK	23:10
Air Blank	0.000	23:11
Control Test	0.076	23:11
Air Blank	0.000	23:11
Subject Sample # 1	0.176	23:12
Breath Volume	2.402	23:12
Air Blank	0.000	23:13
Air Blank	0.000	23:14
Subject Sample # 2	0.179	23:15
Breath Volume	2.304	23:15
Air Blank	0.000	23:16
Control Test	0.076	23:16
Air Blank	0.000	23:16
Diagnostics Check	OK	23:16

Data Download Date/Time: 04/27/2019 12:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BECKMAN,KEVIN,E
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: BLACKMAN,GEOFFREY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 12:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:11

SUBJECT'S LAST NAME: HEATH

DATE/TIME OF TEST: 04/19/2019 13:41:18

Diagnostics Check	OK	13:32	
Air Blank	0.000	13:33	
Control Test	0.076	13:33	
Air Blank	0.000	13:34	
Subject Sample	# 1 0.199	13:37	Volume Not Met
Breath Volume	0.031	13:37	
Air Blank	0.000	13:38	
Air Blank	0.000	13:39	
Subject Sample	# 2	13:39	Subject Test Refused
Breath Volume	0.000	13:39	
Air Blank	0.000	13:40	
Control Test	0.075	13:40	
Air Blank	0.000	13:41	
Diagnostics Check	OK	13:41	

Data Download Date/Time: 04/27/2019 12:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BECKMAN,KEVIN,E
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: BECKWITH,TANNER
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 13:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:25

SUBJECT'S LAST NAME: BATTLE

DATE/TIME OF TEST: 04/20/2019 14:51:51

Diagnostics Check	OK	14:45
Air Blank	0.000	14:46
Control Test	0.078	14:46
Air Blank	0.000	14:47
Subject Sample # 1	0.000	14:47
Breath Volume	1.976	14:47
Air Blank	0.000	14:47
Air Blank	0.000	14:49
Subject Sample # 2	0.000	14:50
Breath Volume	2.386	14:50
Air Blank	0.000	14:50
Control Test	0.078	14:51
Air Blank	0.000	14:51
Diagnostics Check	OK	14:51

Data Download Date/Time: 04/27/2019 12:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001076

INSTRUMENT REGISTERED TO: ZEPHYRHILLS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AIKEN,DURRIEL,R
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: AIKEN,DURRIEL
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 04/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: SHAVER

DATE/TIME OF TEST: 04/05/2019 19:05:00

Diagnostics Check	OK	18:58
Air Blank	0.000	18:59
Control Test	0.081	18:59
Air Blank	0.000	19:00
Subject Sample # 1	0.022	19:00
Breath Volume	1.425	19:00
Air Blank	0.000	19:01
Air Blank	0.000	19:03
Subject Sample # 2	0.019	19:03
Breath Volume	4.238	19:03
Air Blank	0.000	19:04
Control Test	0.081	19:04
Air Blank	0.000	19:04
Diagnostics Check	OK	19:05

Data Download Date/Time: 04/29/2019 16:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001076

INSTRUMENT REGISTERED TO: ZEPHYRHILLS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AIKEN,DURRIEL,R
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: SCHMELTER,CONOR
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 04/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: STEPHENS

DATE/TIME OF TEST: 04/17/2019 00:07:08

Diagnostics Check	OK	23:55	
Air Blank	0.000	23:56	
Control Test	0.082	23:56	
Air Blank	0.000	23:57	
Subject Sample	# 1	00:00	No Sample Provided
Breath Volume	0.000	00:00	
Air Blank	0.000	00:00	
Air Blank	0.000	00:02	
Subject Sample	# 2	00:05	No Sample Provided
Breath Volume	0.000	00:05	
Air Blank	0.000	00:06	
Control Test	0.081	00:06	
Air Blank	0.000	00:07	
Diagnostics Check	OK	00:07	

Data Download Date/Time: 04/29/2019 16:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTWRIGHT JR,GARY,W
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: WARD,SHANE
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 03/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: BETTS

DATE/TIME OF TEST: 04/07/2019 00:58:01

Diagnostics Check	OK	00:51
Air Blank	0.000	00:52
Control Test	0.077	00:52
Air Blank	0.000	00:52
Subject Sample # 1	0.222	00:53
Breath Volume	2.460	00:53
Air Blank	0.000	00:54
Air Blank	0.000	00:55
Subject Sample # 2	0.220	00:56
Breath Volume	2.230	00:56
Air Blank	0.000	00:57
Control Test	0.077	00:57
Air Blank	0.000	00:57
Diagnostics Check	OK	00:58

Data Download Date/Time: 04/30/2019 09:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUSE, RONNIE, L
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: BRANDEBERRY, JEREMY
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 03/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: READY

DATE/TIME OF TEST: 04/09/2019 19:43:33

Diagnostics Check	OK	19:35
Air Blank	0.000	19:36
Control Test	0.077	19:36
Air Blank	0.000	19:37
Subject Sample # 1	0.186	19:38
Breath Volume	1.550	19:38
Air Blank	0.000	19:38
Air Blank	0.000	19:40
Subject Sample # 2	0.191	19:42
Breath Volume	1.511	19:42
Air Blank	0.000	19:42
Control Test	0.078	19:42
Air Blank	0.000	19:43
Diagnostics Check	OK	19:43

Data Download Date/Time: 04/30/2019 09:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANDA,ROBERT,W
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: SCHMIDT,LUKE
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 03/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 14:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:52

SUBJECT'S LAST NAME: ANTRACA

DATE/TIME OF TEST: 04/13/2019 15:28:23

Diagnostics Check	OK	15:21
Air Blank	0.000	15:22
Control Test	0.078	15:22
Air Blank	0.000	15:22
Subject Sample # 1	0.311	15:23
Breath Volume	1.722	15:23
Air Blank	0.000	15:24
Air Blank	0.000	15:26
Subject Sample # 2	0.319	15:26
Breath Volume	1.578	15:26
Air Blank	0.000	15:27
Control Test	0.079	15:27
Air Blank	0.000	15:28
Diagnostics Check	OK	15:28

Data Download Date/Time: 04/30/2019 09:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTWRIGHT JR,GARY,W
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: WARD,SHANE
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 03/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: PEREZ JR

DATE/TIME OF TEST: 04/18/2019 02:19:52

Diagnostics Check	OK	02:13
Air Blank	0.000	02:13
Control Test	0.078	02:14
Air Blank	0.000	02:14
Subject Sample # 1	0.093	02:15
Breath Volume	1.992	02:15
Air Blank	0.000	02:15
Air Blank	0.000	02:17
Subject Sample # 2	0.097	02:18
Breath Volume	1.742	02:18
Air Blank	0.000	02:18
Control Test	0.080	02:19
Air Blank	0.000	02:19
Diagnostics Check	OK	02:19

Data Download Date/Time: 04/30/2019 09:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANDA,ROBERT,W
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: SCHMIDT,LUKE
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 03/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 13:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:55

SUBJECT'S LAST NAME: CONTRERAS ORTIZ

DATE/TIME OF TEST: 04/28/2019 14:38:12

Diagnostics Check	OK	14:31
Air Blank	0.000	14:32
Control Test	0.077	14:32
Air Blank	0.000	14:32
Subject Sample # 1	0.246	14:33
Breath Volume	1.554	14:33
Air Blank	0.000	14:34
Air Blank	0.000	14:36
Subject Sample # 2	0.255	14:36
Breath Volume	1.722	14:36
Air Blank	0.000	14:37
Control Test	0.078	14:37
Air Blank	0.000	14:38
Diagnostics Check	OK	14:38

Data Download Date/Time: 04/30/2019 09:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUSE, RONNIE, L
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: FORT, THOMAS
ARRESTING OFFICER AGENCY: Wauchula PD
LAST AGENCY INSPECTION: 03/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: VARGAS

DATE/TIME OF TEST: 04/28/2019 23:20:22

Diagnostics Check	OK	23:13
Air Blank	0.000	23:14
Control Test	0.078	23:14
Air Blank	0.000	23:14
Subject Sample # 1	0.196	23:15
Breath Volume	1.437	23:15
Air Blank	0.000	23:16
Air Blank	0.000	23:18
Subject Sample # 2	0.198	23:18
Breath Volume	1.722	23:18
Air Blank	0.000	23:19
Control Test	0.077	23:19
Air Blank	0.000	23:20
Diagnostics Check	OK	23:20

Data Download Date/Time: 04/30/2019 09:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUSE, RONNIE, L
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: WARD, SHANE
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 03/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: VELEY

DATE/TIME OF TEST: 04/29/2019 02:42:11

Diagnostics Check	OK	02:32	
Air Blank	0.000	02:32	
Control Test	0.078	02:33	
Air Blank	0.000	02:33	
Subject Sample	# 1	02:36	No Sample Provided
Breath Volume	0.000	02:36	
Air Blank	0.000	02:37	
Air Blank	0.000	02:39	
Subject Sample	# 2	02:40	Subject Test Refused
Breath Volume	0.000	02:40	
Air Blank	0.000	02:41	
Control Test	0.078	02:41	
Air Blank	0.000	02:42	
Diagnostics Check	OK	02:42	

Data Download Date/Time: 04/30/2019 09:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HASTINGS,KENNETH,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: HASTINGS,KENNETH

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 03/12/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17618080A3

DRY GAS EXPIRY DATE: 08/05/2020

VIOLATION CODE: DUI

ARREST TIME: 02:10

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: SYNOVEC

DATE/TIME OF TEST: 04/05/2019 03:10:40

Diagnostics Check	OK	03:03
Air Blank	0.000	03:03
Control Test	0.080	03:04
Air Blank	0.000	03:04
Subject Sample # 1	0.192	03:06
Breath Volume	3.542	03:06
Air Blank	0.000	03:06
Air Blank	0.000	03:08
Subject Sample # 2	0.186	03:09
Breath Volume	2.523	03:09
Air Blank	0.000	03:09
Control Test	0.081	03:10
Air Blank	0.000	03:10
Diagnostics Check	OK	03:10

Data Download Date/Time: 04/15/2019 14:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HASTINGS,KENNETH,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: HASTINGS,KENNETH

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 03/12/2019

Date or Time Changed: Yes

DRY GAS STANDARD LOT #: 17618080A3

DRY GAS EXPIRY DATE: 08/05/2020

VIOLATION CODE: DUI

ARREST TIME: 21:01

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: COSTELO

DATE/TIME OF TEST: 04/05/2019 21:54:53

Diagnostics Check	OK	21:44	
Air Blank	0.000	21:44	
Control Test	0.081	21:45	
Air Blank	0.000	21:45	
Subject Sample	# 1 0.116	21:48	Volume Not Met
Breath Volume	0.113	21:48	
Air Blank	0.000	21:49	
Air Blank	0.000	21:51	
Subject Sample	# 2	21:54	No Sample Provided
Breath Volume	0.000	21:54	
Air Blank		21:54	RFI Detect
Air Blank		21:54	RFI Detect

Data Download Date/Time: 04/15/2019 14:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KATZ,JUSTIN,T
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: KATZ,JUSTIN
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 03/12/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: PALLOTTO

DATE/TIME OF TEST: 04/07/2019 03:43:53

Diagnostics Check	OK	03:32	
Air Blank	0.000	03:32	
Control Test	0.081	03:33	
Air Blank	0.000	03:33	
Subject Sample	# 1	03:36	No Sample Provided
Breath Volume	0.000	03:36	
Air Blank	0.000	03:37	
Air Blank	0.000	03:39	
Subject Sample	# 2	0.201	03:42
Breath Volume	0.062	03:42	Volume Not Met
Air Blank	0.000	03:42	
Control Test	0.081	03:43	
Air Blank	0.000	03:43	
Diagnostics Check	OK	03:43	

Data Download Date/Time: 04/15/2019 14:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KATZ,JUSTIN,T
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: KATZ,JUSTIN
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 03/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: PALLOTTO

DATE/TIME OF TEST: 04/07/2019 03:57:49

Diagnostics Check	OK	03:46	
Air Blank	0.000	03:46	
Control Test	0.080	03:47	
Air Blank	0.000	03:47	
Subject Sample	# 1 0.191	03:50	Volume Not Met
Breath Volume	0.371	03:50	
Air Blank	0.000	03:51	
Air Blank	0.000	03:53	
Subject Sample	# 2 0.181	03:56	Volume Not Met
Breath Volume	0.402	03:56	
Air Blank	0.000	03:56	
Control Test	0.079	03:57	
Air Blank	0.000	03:57	
Diagnostics Check	OK	03:57	

Data Download Date/Time: 04/15/2019 14:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 03/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: TRIPP

DATE/TIME OF TEST: 04/11/2019 04:07:55

Diagnostics Check	OK	03:59
Air Blank	0.000	04:00
Control Test	0.081	04:00
Air Blank	0.000	04:01
Subject Sample # 1	0.121	04:03
Breath Volume	1.789	04:03
Air Blank	0.000	04:04
Air Blank	0.000	04:06
Subject Sample # 2	0.125	04:06
Breath Volume	3.171	04:06
Air Blank	0.000	04:07
Control Test	0.080	04:07
Air Blank	0.000	04:07
Diagnostics Check	OK	04:07

Data Download Date/Time: 04/15/2019 14:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUNT,JERRY,D
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: VANGELLI,JONATHAN
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 03/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: ARNOLD

DATE/TIME OF TEST: 04/13/2019 00:35:32

Diagnostics Check	OK	00:24	
Air Blank	0.000	00:24	
Control Test	0.081	00:24	
Air Blank	0.000	00:25	
Subject Sample	# 1	00:28	No Sample Provided
Breath Volume	0.000	00:28	
Air Blank	0.000	00:29	
Air Blank	0.000	00:31	
Subject Sample	# 2	00:34	No Sample Provided
Breath Volume	0.000	00:34	
Air Blank	0.000	00:34	
Control Test	0.081	00:34	
Air Blank	0.000	00:35	
Diagnostics Check	OK	00:35	

Data Download Date/Time: 04/15/2019 14:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HASTINGS,KENNETH,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: HASTINGS,KENNETH

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 03/12/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17618080A3

DRY GAS EXPIRY DATE: 08/05/2020

VIOLATION CODE: DUI

ARREST TIME: 00:36

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: HAMBROCK

DATE/TIME OF TEST: 04/13/2019 01:22:00

Diagnostics Check	OK	01:14
Air Blank	0.000	01:14
Control Test	0.080	01:15
Air Blank	0.000	01:15
Subject Sample # 1	0.199	01:16
Breath Volume	1.164	01:16
Air Blank	0.000	01:17
Air Blank	0.000	01:19
Subject Sample # 2	0.215	01:20
Breath Volume	1.238	01:20
Air Blank	0.000	01:21
Control Test	0.080	01:21
Air Blank	0.000	01:21
Diagnostics Check	OK	01:22

Data Download Date/Time: 04/15/2019 14:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HASTINGS,KENNETH,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: HASTINGS,KENNETH

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 03/12/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17618080A3

DRY GAS EXPIRY DATE: 08/05/2020

VIOLATION CODE: DUI

ARREST TIME: 03:25

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: HAYNES

DATE/TIME OF TEST: 04/13/2019 04:08:43

Diagnostics Check	OK	04:02
Air Blank	0.000	04:02
Control Test	0.081	04:03
Air Blank	0.000	04:03
Subject Sample # 1	0.069	04:04
Breath Volume	2.144	04:04
Air Blank	0.000	04:04
Air Blank	0.000	04:06
Subject Sample # 2	0.070	04:07
Breath Volume	2.355	04:07
Air Blank	0.000	04:07
Control Test	0.080	04:08
Air Blank	0.000	04:08
Diagnostics Check	OK	04:08

Data Download Date/Time: 04/15/2019 14:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KATZ,JUSTIN,T
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: KWIATKOWSKA,SYLWIA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 03/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:46

SUBJECT'S LAST NAME: PERKINS

DATE/TIME OF TEST: 04/14/2019 18:14:04

Diagnostics Check	OK	18:07
Air Blank	0.000	18:08
Control Test	0.081	18:08
Air Blank	0.000	18:09
Subject Sample # 1	0.233	18:09
Breath Volume	1.601	18:09
Air Blank	0.000	18:10
Air Blank	0.000	18:12
Subject Sample # 2	0.234	18:12
Breath Volume	1.468	18:12
Air Blank	0.000	18:13
Control Test	0.080	18:13
Air Blank	0.000	18:13
Diagnostics Check	OK	18:14

Data Download Date/Time: 04/15/2019 14:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: SILVERSTEIN,CHRISTINA

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 04/15/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17618080A3

DRY GAS EXPIRY DATE: 08/05/2020

VIOLATION CODE: DUI

ARREST TIME: 03:14

GENDER: Female

OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: PERNIA LINARES

DATE/TIME OF TEST: 04/20/2019 04:03:46

Diagnostics Check	OK	03:56
Air Blank	0.000	03:57
Control Test	0.081	03:57
Air Blank	0.000	03:58
Subject Sample # 1	0.131	03:59
Breath Volume	2.757	03:59
Air Blank	0.000	03:59
Air Blank	0.000	04:01
Subject Sample # 2	0.131	04:02
Breath Volume	2.382	04:02
Air Blank	0.000	04:02
Control Test	0.080	04:03
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time: 05/21/2019 14:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HASTINGS,KENNETH,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: HASTINGS,KENNETH

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 04/15/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17618080A3

DRY GAS EXPIRY DATE: 08/05/2020

VIOLATION CODE: DUI

ARREST TIME: 04:04

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: HATIMI

DATE/TIME OF TEST: 04/20/2019 04:47:38

Diagnostics Check	OK	04:43	
Air Blank	0.000	04:44	
Control Test	0.081	04:44	
Air Blank	0.000	04:44	
Subject Sample	# 1	04:46	Subject Test Refused
Breath Volume	0.000	04:46	
Air Blank	0.000	04:46	
Control Test	0.080	04:47	
Air Blank	0.000	04:47	
Diagnostics Check	OK	04:47	

Data Download Date/Time: 05/21/2019 14:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: SILVERSTEIN,CHRISTINA

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 04/15/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17618080A3

DRY GAS EXPIRY DATE: 08/05/2020

VIOLATION CODE: DUI

ARREST TIME: 23:37

GENDER: Female

OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: FLEMING

DATE/TIME OF TEST: 04/21/2019 00:23:10

Diagnostics Check	OK	00:16
Air Blank	0.000	00:17
Control Test	0.081	00:17
Air Blank	0.000	00:17
Subject Sample # 1	0.141	00:18
Breath Volume	1.207	00:18
Air Blank	0.000	00:19
Air Blank	0.000	00:21
Subject Sample # 2	0.160	00:21
Breath Volume	1.214	00:21
Air Blank	0.000	00:22
Control Test	0.080	00:22
Air Blank	0.000	00:23
Diagnostics Check	OK	00:23

Data Download Date/Time: 05/21/2019 14:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 04/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: BETZ

DATE/TIME OF TEST: 04/25/2019 02:59:53

Diagnostics Check	OK	02:53
Air Blank	0.000	02:54
Control Test	0.080	02:54
Air Blank	0.000	02:54
Subject Sample # 1	0.140	02:55
Breath Volume	3.062	02:55
Air Blank	0.000	02:55
Air Blank	0.000	02:57
Subject Sample # 2	0.133	02:58
Breath Volume	2.453	02:58
Air Blank	0.000	02:58
Control Test	0.080	02:59
Air Blank	0.000	02:59
Diagnostics Check	OK	02:59

Data Download Date/Time: 05/21/2019 14:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CITRANO, ANTHONY,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: CITRANO, ANTHONY

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 04/15/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17618080A3

DRY GAS EXPIRY DATE: 08/05/2020

VIOLATION CODE: DUI

ARREST TIME: 07:20

GENDER: Male

OBSERVATION PERIOD BEGAN: 07:44

SUBJECT'S LAST NAME: FREUND

DATE/TIME OF TEST: 04/25/2019 08:15:52

Diagnostics Check	OK	08:09
Air Blank	0.000	08:10
Control Test	0.081	08:10
Air Blank	0.000	08:11
Subject Sample # 1	0.000	08:11
Breath Volume	3.167	08:11
Air Blank	0.000	08:12
Air Blank	0.000	08:14
Subject Sample # 2	0.000	08:14
Breath Volume	2.589	08:14
Air Blank	0.000	08:14
Control Test	0.080	08:15
Air Blank	0.000	08:15
Diagnostics Check	OK	08:15

Data Download Date/Time: 05/21/2019 14:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: SILVERSTEIN,CHRISTINA

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 04/15/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17618080A3

DRY GAS EXPIRY DATE: 08/05/2020

VIOLATION CODE: DUI

ARREST TIME: 03:41

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: MILEY

DATE/TIME OF TEST: 04/26/2019 04:22:58

Diagnostics Check	OK	04:16
Air Blank	0.000	04:17
Control Test	0.081	04:17
Air Blank	0.000	04:17
Subject Sample # 1	0.197	04:18
Breath Volume	2.734	04:18
Air Blank	0.000	04:19
Air Blank	0.000	04:21
Subject Sample # 2	0.186	04:21
Breath Volume	1.652	04:21
Air Blank	0.000	04:22
Control Test	0.080	04:22
Air Blank	0.000	04:22
Diagnostics Check	OK	04:22

Data Download Date/Time: 05/21/2019 14:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HASTINGS,KENNETH,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: HASTINGS,KENNETH

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 04/15/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17618080A3

DRY GAS EXPIRY DATE: 08/05/2020

VIOLATION CODE: DUI

ARREST TIME: 21:12

GENDER: Female

OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: ARGENAL-HERRERA

DATE/TIME OF TEST: 04/26/2019 22:07:52

Diagnostics Check	OK	22:01
Air Blank	0.000	22:01
Control Test	0.080	22:02
Air Blank	0.000	22:02
Subject Sample # 1	0.000	22:03
Breath Volume	2.257	22:03
Air Blank	0.000	22:03
Air Blank	0.000	22:05
Subject Sample # 2	0.000	22:06
Breath Volume	2.050	22:06
Air Blank	0.000	22:06
Control Test	0.080	22:07
Air Blank	0.000	22:07
Diagnostics Check	OK	22:07

Data Download Date/Time: 05/21/2019 14:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ALAMEDA,JORGE,L

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: ALAMEDA,JORGE

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 04/15/2019

Date or Time Changed: Yes

DRY GAS STANDARD LOT #: 17618080A3

DRY GAS EXPIRY DATE: 08/05/2020

VIOLATION CODE: DUI

ARREST TIME: 17:17

GENDER: Female

OBSERVATION PERIOD BEGAN: 18:04

SUBJECT'S LAST NAME: SMADES

DATE/TIME OF TEST: 04/30/2019 18:28:41

Diagnostics Check	OK	18:25	
Air Blank	0.000	18:25	
Control Test	0.081	18:26	
Air Blank	0.000	18:26	
Subject Sample	# 1	18:27	Subject Test Refused
Breath Volume	0.000	18:27	
Air Blank	0.000	18:27	
Control Test	0.080	18:28	
Air Blank	0.000	18:28	
Diagnostics Check	OK	18:28	

Data Download Date/Time: 05/21/2019 14:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JEUDY,RINALDO,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HAMILTON,ROBERT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 04/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: MUNIZ ORTEGA

DATE/TIME OF TEST: 04/28/2019 22:37:07

Diagnostics Check	OK	22:26	
Air Blank	0.000	22:26	
Control Test	0.076	22:26	
Air Blank	0.000	22:27	
Subject Sample	# 1 0.092	22:30	Volume Not Met
Breath Volume	0.035	22:30	
Air Blank	0.000	22:31	
Air Blank	0.000	22:33	
Subject Sample	# 2	22:35	Subject Test Refused
Breath Volume	0.023	22:35	
Air Blank	0.000	22:36	
Control Test	0.076	22:36	
Air Blank	0.000	22:37	
Diagnostics Check	OK	22:37	

Data Download Date/Time: 04/30/2019 16:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHOUINARD,JEFFREY,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: CHOUINARD,JEFFREY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 11:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: TRAVERS

DATE/TIME OF TEST: 04/29/2019 01:30:23

Diagnostics Check	OK	01:23
Air Blank	0.000	01:23
Control Test	0.077	01:24
Air Blank	0.000	01:24
Subject Sample # 1	0.136	01:25
Breath Volume	4.589	01:25
Air Blank	0.000	01:26
Air Blank	0.000	01:27
Subject Sample # 2	0.134	01:28
Breath Volume	4.441	01:28
Air Blank	0.000	01:29
Control Test	0.075	01:29
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

Data Download Date/Time: 04/30/2019 16:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,ROBERT,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: DEMARK,HAROLD
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: HANCOCK

DATE/TIME OF TEST: 04/30/2019 04:02:24

Diagnostics Check	OK	03:59	
Air Blank	0.000	04:00	
Control Test	0.078	04:00	
Air Blank	0.000	04:00	
Subject Sample	# 1	04:01	Subject Test Refused
Breath Volume	0.000	04:01	
Air Blank	0.000	04:01	
Control Test	0.078	04:01	
Air Blank	0.000	04:02	
Diagnostics Check	OK	04:02	

Data Download Date/Time: 04/30/2019 16:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANNON, MATTHEW, F
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HANNON, MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 06:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:30

SUBJECT'S LAST NAME: ZIO

DATE/TIME OF TEST: 04/01/2019 09:06:45

Diagnostics Check	OK	08:58	
Air Blank	0.000	08:58	
Control Test	0.081	08:59	
Air Blank	0.000	08:59	
Subject Sample	# 1	09:02	No Sample Provided
Breath Volume	0.000	09:02	
Air Blank	0.000	09:03	
Air Blank	0.000	09:05	
Subject Sample	# 2	09:05	Subject Test Refused
Breath Volume	0.000	09:05	
Air Blank	0.000	09:05	
Control Test	0.079	09:06	
Air Blank	0.000	09:06	
Diagnostics Check	OK	09:06	

Data Download Date/Time: 04/27/2019 11:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUNNER,JOSEPH,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: RUNNER,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: REESE

DATE/TIME OF TEST: 04/02/2019 00:08:40

Diagnostics Check	OK	00:07	
Air Blank	0.000	00:07	
Control Test	0.079	00:08	
Air Blank		00:08	RFI Detect
Air Blank		00:08	RFI Detect

Data Download Date/Time: 04/27/2019 11:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUNNER,JOSEPH,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: RUNNER,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: REESE

DATE/TIME OF TEST: 04/02/2019 00:23:24

Diagnostics Check	OK	00:22	
Air Blank		00:23	RFI Detect
Air Blank		00:23	RFI Detect

Data Download Date/Time: 04/27/2019 11:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSENBLOOM,MATTHEW,S
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: HAMMELMAN,TYLER
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: MACUMBER

DATE/TIME OF TEST: 04/02/2019 03:36:31

Diagnostics Check	OK	03:29
Air Blank	0.000	03:30
Control Test	0.080	03:30
Air Blank	0.000	03:31
Subject Sample # 1	0.159	03:31
Breath Volume	2.402	03:31
Air Blank	0.000	03:32
Air Blank	0.000	03:34
Subject Sample # 2	0.160	03:35
Breath Volume	2.250	03:35
Air Blank	0.000	03:35
Control Test	0.081	03:35
Air Blank	0.000	03:36
Diagnostics Check	OK	03:36

Data Download Date/Time: 04/27/2019 11:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARROW,M,A
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: BOLES,P
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 08:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:41

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 04/04/2019 10:18:13

Diagnostics Check	OK	10:12
Air Blank	0.000	10:12
Control Test	0.081	10:12
Air Blank	0.000	10:13
Subject Sample # 1	0.168	10:13
Breath Volume	3.574	10:13
Air Blank	0.000	10:14
Air Blank	0.000	10:16
Subject Sample # 2	0.169	10:16
Breath Volume	3.574	10:16
Air Blank	0.000	10:17
Control Test	0.081	10:17
Air Blank	0.000	10:18
Diagnostics Check	OK	10:18

Data Download Date/Time: 04/27/2019 12:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNORR,PETER,J
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: KNORR,PETER
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 13:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:55

SUBJECT'S LAST NAME: SNELL

DATE/TIME OF TEST: 04/04/2019 14:29:00

Diagnostics Check	OK	14:15	
Air Blank	0.000	14:16	
Control Test	0.080	14:16	
Air Blank	0.000	14:16	
Subject Sample	# 1 0.300	14:17	No .020 agreement
Breath Volume	1.265	14:17	
Air Blank	0.000	14:18	
Air Blank	0.000	14:20	
Subject Sample	# 2 0.226	14:23	Volume Not Met
Breath Volume	0.074	14:23	
Air Blank	0.000	14:24	
Air Blank	0.000	14:25	
Subject Sample	# 3 0.262	14:27	No .020 agreement
Breath Volume	1.312	14:27	
Air Blank	0.000	14:28	
Control Test	0.078	14:28	
Air Blank	0.000	14:28	
Diagnostics Check	OK	14:29	

Data Download Date/Time: 04/27/2019 12:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAUGHERTY,D,M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: DAUGHERTY,D
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:51

SUBJECT'S LAST NAME: ROSARIO

DATE/TIME OF TEST: 04/04/2019 20:22:22

Diagnostics Check	OK	20:12	
Air Blank	0.000	20:12	
Control Test	0.081	20:12	
Air Blank	0.000	20:13	
Subject Sample	# 1 0.125	20:16	Volume Not Met
Breath Volume	0.511	20:16	
Air Blank	0.000	20:17	
Air Blank	0.000	20:19	
Subject Sample	# 2	20:20	Subject Test Refused
Breath Volume	0.824	20:20	
Air Blank	0.000	20:21	
Control Test	0.080	20:21	
Air Blank	0.000	20:22	
Diagnostics Check	OK	20:22	

Data Download Date/Time: 04/27/2019 12:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK, TYLER, J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KIRK, TYLER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: HUDDALL

DATE/TIME OF TEST: 04/05/2019 23:38:39

Diagnostics Check	OK	23:32
Air Blank	0.000	23:32
Control Test	0.078	23:33
Air Blank	0.000	23:33
Subject Sample # 1	0.192	23:34
Breath Volume	2.730	23:34
Air Blank	0.000	23:34
Air Blank	0.000	23:36
Subject Sample # 2	0.191	23:37
Breath Volume	2.363	23:37
Air Blank	0.000	23:37
Control Test	0.080	23:38
Air Blank	0.000	23:38
Diagnostics Check	OK	23:38

Data Download Date/Time: 04/27/2019 12:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUNNER,JOSEPH,A
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: ALLMAN,WAYNE
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 04/06/2019 04:55:51

Diagnostics Check	OK	04:48
Air Blank	0.000	04:49
Control Test	0.077	04:49
Air Blank	0.000	04:50
Subject Sample # 1	0.195	04:51
Breath Volume	1.335	04:51
Air Blank	0.000	04:51
Air Blank	0.000	04:53
Subject Sample # 2	0.199	04:54
Breath Volume	1.867	04:54
Air Blank	0.000	04:54
Control Test	0.080	04:55
Air Blank	0.000	04:55
Diagnostics Check	OK	04:55

Data Download Date/Time: 04/27/2019 12:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANE,MARC,T
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: LANE,MARC
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 14:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:26

SUBJECT'S LAST NAME: SCHLEPPI

DATE/TIME OF TEST: 04/10/2019 15:52:22

Diagnostics Check	OK	15:46
Air Blank	0.000	15:46
Control Test	0.081	15:47
Air Blank	0.000	15:47
Subject Sample # 1	0.000	15:48
Breath Volume	1.109	15:48
Air Blank	0.000	15:48
Air Blank	0.000	15:50
Subject Sample # 2	0.000	15:51
Breath Volume	1.164	15:51
Air Blank	0.000	15:51
Control Test	0.081	15:51
Air Blank	0.000	15:52
Diagnostics Check	OK	15:52

Data Download Date/Time: 04/27/2019 12:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANNON, MATTHEW, F
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HANNON, MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: GEORGE

DATE/TIME OF TEST: 04/13/2019 03:54:56

Diagnostics Check	OK	03:46	
Air Blank	0.000	03:46	
Control Test	0.081	03:47	
Air Blank	0.000	03:47	
Subject Sample	# 1	03:50	No Sample Provided
Breath Volume	0.000	03:50	
Air Blank	0.000	03:51	
Air Blank	0.000	03:53	
Subject Sample	# 2	03:53	Subject Test Refused
Breath Volume	0.000	03:53	
Air Blank	0.000	03:54	
Control Test	0.080	03:54	
Air Blank	0.000	03:54	
Diagnostics Check	OK	03:54	

Data Download Date/Time: 04/27/2019 12:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANNON, MATTHEW, F
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HANNON, MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: SNYDER

DATE/TIME OF TEST: 04/14/2019 02:25:26

Diagnostics Check	OK	02:17	
Air Blank	0.000	02:17	
Control Test	0.081	02:17	
Air Blank	0.000	02:18	
Subject Sample	# 1 0.083	02:21	Volume Not Met
Breath Volume	0.199	02:21	
Air Blank	0.000	02:22	
Air Blank	0.000	02:23	
Subject Sample	# 2	02:24	Subject Test Refused
Breath Volume	0.117	02:24	
Air Blank		02:24	RFI Detect
Air Blank	0.000	02:25	

Data Download Date/Time: 04/27/2019 12:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANNON, MATTHEW, F
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HANNON, MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: SNYDER

DATE/TIME OF TEST: 04/14/2019 02:35:24

Diagnostics Check	OK	02:26	
Air Blank	0.000	02:27	
Control Test	0.081	02:27	
Air Blank	0.000	02:28	
Subject Sample	# 1	02:31	No Sample Provided
Breath Volume	0.000	02:31	
Air Blank	0.000	02:31	
Air Blank	0.000	02:33	
Subject Sample	# 2	02:34	Subject Test Refused
Breath Volume	0.000	02:34	
Air Blank	0.000	02:34	
Control Test	0.081	02:34	
Air Blank	0.000	02:35	
Diagnostics Check	OK	02:35	

Data Download Date/Time: 04/27/2019 12:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,RAEFORD,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GRIFFIN,RAEFORD
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 04/17/2019 00:51:29

Diagnostics Check	OK	00:43	
Air Blank	0.000	00:43	
Control Test	0.081	00:43	
Air Blank	0.000	00:44	
Subject Sample	# 1	00:47	No Sample Provided
Breath Volume	0.000	00:47	
Air Blank	0.000	00:47	
Air Blank	0.000	00:49	
Subject Sample	# 2	00:50	Subject Test Refused
Breath Volume	0.000	00:50	
Air Blank	0.000	00:50	
Control Test	0.080	00:50	
Air Blank	0.000	00:51	
Diagnostics Check	OK	00:51	

Data Download Date/Time: 04/27/2019 12:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANNON, MATTHEW, F
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HANNON, MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: FANTOLI

DATE/TIME OF TEST: 04/17/2019 23:28:09

Diagnostics Check	OK	23:21
Air Blank	0.000	23:22
Control Test	0.081	23:22
Air Blank	0.000	23:22
Subject Sample # 1	0.000	23:23
Breath Volume	1.148	23:23
Air Blank	0.000	23:24
Air Blank	0.000	23:26
Subject Sample # 2	0.000	23:26
Breath Volume	1.316	23:26
Air Blank	0.000	23:27
Control Test	0.080	23:27
Air Blank	0.000	23:28
Diagnostics Check	OK	23:28

Data Download Date/Time: 04/27/2019 12:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLY, WESLEY, R
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KELLY, WESLEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 10:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:15

SUBJECT'S LAST NAME: HERRERA

DATE/TIME OF TEST: 04/18/2019 11:43:11

Diagnostics Check	OK	11:36
Air Blank	0.000	11:37
Control Test	0.081	11:37
Air Blank	0.000	11:37
Subject Sample # 1	0.000	11:38
Breath Volume	2.000	11:38
Air Blank	0.000	11:39
Air Blank	0.000	11:41
Subject Sample # 2	0.000	11:41
Breath Volume	1.789	11:41
Air Blank	0.000	11:42
Control Test	0.081	11:42
Air Blank	0.000	11:43
Diagnostics Check	OK	11:43

Data Download Date/Time: 04/27/2019 12:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANNON,MATTHEW,F
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HANNON,MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: PECORA

DATE/TIME OF TEST: 04/19/2019 02:08:49

Diagnostics Check	OK	02:02
Air Blank	0.000	02:03
Control Test	0.081	02:03
Air Blank	0.000	02:03
Subject Sample # 1	0.137	02:04
Breath Volume	2.898	02:04
Air Blank	0.000	02:05
Air Blank	0.000	02:06
Subject Sample # 2	0.123	02:07
Breath Volume	2.066	02:07
Air Blank	0.000	02:07
Control Test	0.079	02:08
Air Blank	0.000	02:08
Diagnostics Check	OK	02:08

Data Download Date/Time: 04/27/2019 12:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COFFEE,JASON,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: COFFEE,JASON
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: HEINKEL

DATE/TIME OF TEST: 04/19/2019 04:06:05

Diagnostics Check	OK	03:59
Air Blank	0.000	03:59
Control Test	0.080	04:00
Air Blank	0.000	04:00
Subject Sample # 1	0.212	04:01
Breath Volume	1.984	04:01
Air Blank	0.000	04:02
Air Blank	0.000	04:03
Subject Sample # 2	0.202	04:04
Breath Volume	1.843	04:04
Air Blank	0.000	04:05
Control Test	0.079	04:05
Air Blank	0.000	04:06
Diagnostics Check	OK	04:06

Data Download Date/Time: 04/27/2019 12:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,BARRY,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: NIXON,BARRY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: SCHNEIDER

DATE/TIME OF TEST: 04/20/2019 00:15:08

Diagnostics Check	OK	00:09
Air Blank	0.000	00:09
Control Test	0.080	00:09
Air Blank	0.000	00:10
Subject Sample # 1	0.211	00:10
Breath Volume	1.457	00:10
Air Blank	0.000	00:11
Air Blank	0.000	00:13
Subject Sample # 2	0.224	00:13
Breath Volume	1.242	00:13
Air Blank	0.000	00:14
Control Test	0.080	00:14
Air Blank	0.000	00:15
Diagnostics Check	OK	00:15

Data Download Date/Time: 04/27/2019 12:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAUGHERTY,D,M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: DAUGHERTY,D
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: PICKERING

DATE/TIME OF TEST: 04/20/2019 02:39:03

Diagnostics Check	OK	02:32
Air Blank	0.000	02:32
Control Test	0.080	02:33
Air Blank	0.000	02:33
Subject Sample # 1	0.242	02:34
Breath Volume	1.339	02:34
Air Blank	0.000	02:34
Air Blank	0.000	02:36
Subject Sample # 2	0.241	02:37
Breath Volume	1.125	02:37
Air Blank	0.000	02:38
Control Test	0.081	02:38
Air Blank	0.000	02:38
Diagnostics Check	OK	02:39

Data Download Date/Time: 04/27/2019 12:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, WILLIAM, T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: MCGEE

DATE/TIME OF TEST: 04/20/2019 03:22:52

Diagnostics Check	OK	03:16
Air Blank	0.000	03:16
Control Test	0.080	03:16
Air Blank	0.000	03:17
Subject Sample # 1	0.155	03:18
Breath Volume	1.515	03:18
Air Blank	0.000	03:19
Air Blank	0.000	03:20
Subject Sample # 2	0.175	03:21
Breath Volume	1.269	03:21
Air Blank	0.000	03:21
Control Test	0.080	03:22
Air Blank	0.000	03:22
Diagnostics Check	OK	03:22

Data Download Date/Time: 04/27/2019 12:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,RAEFORD,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GRIFFIN,RAEFORD
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:40

SUBJECT'S LAST NAME: MCCASLIN

DATE/TIME OF TEST: 04/20/2019 18:31:31

Diagnostics Check	OK	18:22	
Air Blank	0.000	18:22	
Control Test	0.082	18:22	
Air Blank	0.000	18:23	
Subject Sample	# 1	18:26	No Sample Provided
Breath Volume	0.000	18:26	
Air Blank	0.000	18:26	
Air Blank	0.000	18:28	
Subject Sample	# 2	18:30	Subject Test Refused
Breath Volume	0.132	18:30	
Air Blank	0.000	18:30	
Control Test	0.078	18:30	
Air Blank	0.000	18:31	
Diagnostics Check	OK	18:31	

Data Download Date/Time: 04/27/2019 12:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEQUEIRA,NATHAN,C
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SEQUEIRA,NATHAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 06:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: BIGGANS

DATE/TIME OF TEST: 04/20/2019 20:07:13

Diagnostics Check	OK	20:00
Air Blank	0.000	20:00
Control Test	0.081	20:01
Air Blank	0.000	20:01
Subject Sample # 1	0.043	20:02
Breath Volume	1.699	20:02
Air Blank	0.000	20:03
Air Blank	0.000	20:05
Subject Sample # 2	0.040	20:05
Breath Volume	2.433	20:05
Air Blank	0.000	20:06
Control Test	0.081	20:06
Air Blank	0.000	20:07
Diagnostics Check	OK	20:07

Data Download Date/Time: 04/27/2019 12:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUNNER,JOSEPH,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: RUNNER,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: KRYSHESKI

DATE/TIME OF TEST: 04/21/2019 03:08:43

Diagnostics Check	OK	03:02
Air Blank	0.000	03:02
Control Test	0.080	03:02
Air Blank	0.000	03:03
Subject Sample # 1	0.159	03:04
Breath Volume	1.437	03:04
Air Blank	0.000	03:04
Air Blank	0.000	03:06
Subject Sample # 2	0.160	03:07
Breath Volume	1.515	03:07
Air Blank	0.000	03:07
Control Test	0.080	03:08
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

Data Download Date/Time: 04/27/2019 12:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUCHANAN,SUMMER,N
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BUCHANAN,SUMMER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: BUCKLEN

DATE/TIME OF TEST: 04/21/2019 04:02:59

Diagnostics Check	OK	03:56
Air Blank	0.000	03:56
Control Test	0.080	03:56
Air Blank	0.000	03:57
Subject Sample # 1	0.191	03:58
Breath Volume	3.300	03:58
Air Blank	0.000	03:59
Air Blank	0.000	04:00
Subject Sample # 2	0.198	04:01
Breath Volume	3.015	04:01
Air Blank	0.000	04:02
Control Test	0.080	04:02
Air Blank	0.000	04:02
Diagnostics Check	OK	04:02

Data Download Date/Time: 04/27/2019 12:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLY, WESLEY, R
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KELLY, WESLEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 08:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:25

SUBJECT'S LAST NAME: BAXTER

DATE/TIME OF TEST: 04/22/2019 08:53:15

Diagnostics Check	OK	08:46
Air Blank	0.000	08:46
Control Test	0.082	08:47
Air Blank	0.000	08:47
Subject Sample # 1	0.000	08:48
Breath Volume	2.375	08:48
Air Blank	0.000	08:49
Air Blank	0.000	08:51
Subject Sample # 2	0.000	08:51
Breath Volume	1.445	08:51
Air Blank	0.000	08:52
Control Test	0.081	08:52
Air Blank	0.000	08:53
Diagnostics Check	OK	08:53

Data Download Date/Time: 04/27/2019 12:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLY, WESLEY, R
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KELLY, WESLEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: BERRY

DATE/TIME OF TEST: 04/22/2019 16:58:50

Diagnostics Check	OK	16:51
Air Blank	0.000	16:52
Control Test	0.081	16:52
Air Blank	0.000	16:53
Subject Sample # 1	0.000	16:54
Breath Volume	2.398	16:54
Air Blank	0.000	16:54
Air Blank	0.000	16:56
Subject Sample # 2	0.000	16:57
Breath Volume	2.125	16:57
Air Blank	0.000	16:57
Control Test	0.080	16:58
Air Blank	0.000	16:58
Diagnostics Check	OK	16:58

Data Download Date/Time: 04/27/2019 12:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,RAEFORD,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HAMMELMAN,TYLER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: LITTS

DATE/TIME OF TEST: 04/26/2019 04:23:47

Diagnostics Check	OK	04:14
Air Blank	0.000	04:15
Control Test	0.081	04:15
Air Blank	0.000	04:15
Subject Sample # 1	0.151	04:16
Breath Volume	2.878	04:16
Air Blank	0.000	04:17
Air Blank	0.000	04:18
Subject Sample # 2	0.128	04:19
Breath Volume	1.648	04:19
Air Blank	0.000	04:20
Air Blank	0.000	04:21
Subject Sample # 3	0.129	04:22
Breath Volume	1.550	04:22
Air Blank	0.000	04:22
Control Test	0.080	04:23
Air Blank	0.000	04:23
Diagnostics Check	OK	04:23

Data Download Date/Time: 04/27/2019 12:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REAY,ALEX,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: REAY,ALEX
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: SZEWCZYK

DATE/TIME OF TEST: 04/26/2019 19:35:34

Diagnostics Check	OK	19:29
Air Blank	0.000	19:29
Control Test	0.081	19:29
Air Blank	0.000	19:30
Subject Sample # 1	0.087	19:30
Breath Volume	1.781	19:30
Air Blank	0.000	19:31
Air Blank	0.000	19:33
Subject Sample # 2	0.088	19:34
Breath Volume	2.386	19:34
Air Blank	0.000	19:34
Control Test	0.081	19:34
Air Blank	0.000	19:35
Diagnostics Check	OK	19:35

Data Download Date/Time: 04/27/2019 12:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANNON, MATTHEW, F
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HANNON, MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 04/28/2019 22:00:23

Diagnostics Check	OK	21:54
Air Blank	0.000	21:54
Control Test	0.078	21:54
Air Blank	0.000	21:55
Subject Sample # 1	0.292	21:55
Breath Volume	3.589	21:55
Air Blank	0.000	21:56
Air Blank	0.000	21:58
Subject Sample # 2	0.295	21:58
Breath Volume	3.578	21:58
Air Blank	0.000	21:59
Control Test	0.078	21:59
Air Blank	0.000	22:00
Diagnostics Check	OK	22:00

Data Download Date/Time: 05/16/2019 19:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONNELL,JACOB,B
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: CAMPANA,NICK
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 04/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: MALEY'NASH

DATE/TIME OF TEST: 04/29/2019 00:02:04

Diagnostics Check	OK	23:53
Air Blank	0.000	23:54
Control Test	0.080	23:54
Air Blank	0.000	23:54
Subject Sample # 2	0.203	23:55
Breath Volume	2.390	23:55
Air Blank	0.000	23:56
Air Blank	0.000	23:57
Subject Sample # 1	0.183	00:00
Breath Volume	2.046	00:00
Air Blank	0.000	00:01
Control Test	0.079	00:01
Air Blank	0.000	00:01
Diagnostics Check	OK	00:02

Data Download Date/Time: 05/16/2019 19:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARQUHARSON,KEVIN,
BREATH TEST OPERATOR AGENCY: Plantation PD
ARRESTING OFFICER: FARQUHARSON,KEVIN
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A5
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: MARQUEZ-MORALES

DATE/TIME OF TEST: 04/14/2019 04:40:08

Diagnostics Check	OK	04:31
Air Blank	0.000	04:32
Control Test	0.078	04:32
Air Blank	0.000	04:33
Subject Sample # 1	0.154	04:35
Breath Volume	3.335	04:35
Air Blank	0.000	04:35
Air Blank	0.000	04:37
Subject Sample # 2	0.144	04:38
Breath Volume	2.382	04:38
Air Blank	0.000	04:39
Control Test	0.077	04:39
Air Blank	0.000	04:40
Diagnostics Check	OK	04:40

Data Download Date/Time: 04/17/2019 20:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IASILLO,OFC M,
BREATH TEST OPERATOR AGENCY: Plantation PD
ARRESTING OFFICER: BEAUREGARD,OFC. J
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 04/17/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A5
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:25

SUBJECT'S LAST NAME: O'NEAL

DATE/TIME OF TEST: 04/18/2019 18:52:40

Diagnostics Check	OK	18:45
Air Blank	0.000	18:46
Control Test	0.078	18:46
Air Blank	0.000	18:46
Subject Sample # 1	0.180	18:48
Breath Volume	1.687	18:48
Air Blank	0.000	18:48
Air Blank	0.000	18:50
Subject Sample # 2	0.174	18:51
Breath Volume	2.402	18:51
Air Blank	0.000	18:51
Control Test	0.076	18:52
Air Blank	0.000	18:52
Diagnostics Check	OK	18:52

Data Download Date/Time: 05/29/2019 15:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REIDY,C,
BREATH TEST OPERATOR AGENCY: Plantation PD
ARRESTING OFFICER: GRANT,C
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 04/17/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 10818080A5
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 12:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:53

SUBJECT'S LAST NAME: PISCIONE

DATE/TIME OF TEST: 04/29/2019 13:29:38

Diagnostics Check	OK	13:21
Air Blank	0.000	13:22
Control Test	0.078	13:22
Air Blank	0.000	13:23
Subject Sample # 1	0.168	13:24
Breath Volume	3.382	13:24
Air Blank	0.000	13:25
Air Blank	0.000	13:27
Subject Sample # 2	0.156	13:28
Breath Volume	2.945	13:28
Air Blank	0.000	13:28
Control Test	0.077	13:29
Air Blank	0.000	13:29
Diagnostics Check	OK	13:29

Data Download Date/Time: 05/29/2019 15:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BASS,H,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: STEPHENS

DATE/TIME OF TEST: 04/04/2019 01:20:54

Diagnostics Check	OK	01:18	
Air Blank	0.000	01:18	
Control Test	0.082	01:18	
Air Blank	0.000	01:19	
Subject Sample	# 1	01:19	Subject Test Refused
Breath Volume	0.000	01:19	
Air Blank	0.000	01:19	
Control Test	0.082	01:20	
Air Blank	0.000	01:20	
Diagnostics Check	OK	01:20	

Data Download Date/Time: 04/16/2019 15:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AUGUSTINE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 04/09/2019 23:50:41

Diagnostics Check	OK	23:44
Air Blank	0.000	23:44
Control Test	0.081	23:45
Air Blank	0.000	23:45
Subject Sample # 1	0.354	23:46
Breath Volume	3.707	23:46
Air Blank	0.000	23:46
Air Blank	0.000	23:48
Subject Sample # 2	0.339	23:49
Breath Volume	4.027	23:49
Air Blank	0.000	23:49
Control Test	0.081	23:50
Air Blank	0.000	23:50
Diagnostics Check	OK	23:50

Data Download Date/Time: 04/16/2019 15:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AUGUSTINE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: SHACKA

DATE/TIME OF TEST: 04/10/2019 00:40:27

Diagnostics Check	OK	00:33
Air Blank	0.000	00:34
Control Test	0.080	00:34
Air Blank	0.000	00:35
Subject Sample # 1	0.200	00:35
Breath Volume	2.351	00:35
Air Blank	0.000	00:36
Air Blank	0.000	00:38
Subject Sample # 2	0.201	00:38
Breath Volume	2.406	00:38
Air Blank	0.000	00:39
Control Test	0.081	00:39
Air Blank	0.000	00:40
Diagnostics Check	OK	00:40

Data Download Date/Time: 04/16/2019 15:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAMES,A,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MECHLOWITZ,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 08:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:43

SUBJECT'S LAST NAME: DEVOE

DATE/TIME OF TEST: 04/11/2019 11:10:34

Diagnostics Check	OK	11:04
Air Blank	0.000	11:04
Control Test	0.081	11:04
Air Blank	0.000	11:05
Subject Sample # 1	0.173	11:06
Breath Volume	2.066	11:06
Air Blank	0.000	11:06
Air Blank	0.000	11:08
Subject Sample # 2	0.184	11:09
Breath Volume	2.335	11:09
Air Blank	0.000	11:09
Control Test	0.080	11:09
Air Blank	0.000	11:10
Diagnostics Check	OK	11:10

Data Download Date/Time: 04/16/2019 15:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,D,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DIAZ,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 04/12/2019 02:42:52

Diagnostics Check	OK	02:39	
Air Blank	0.000	02:40	
Control Test	0.081	02:40	
Air Blank	0.000	02:41	
Subject Sample	# 1	02:41	Subject Test Refused
Breath Volume	0.000	02:41	
Air Blank	0.000	02:41	
Control Test	0.081	02:42	
Air Blank	0.000	02:42	
Diagnostics Check	OK	02:42	

Data Download Date/Time: 04/16/2019 15:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,D,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TOKAREK,P
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 08:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:22

SUBJECT'S LAST NAME: GOMEZ LOPEZ

DATE/TIME OF TEST: 04/14/2019 10:51:31

Diagnostics Check	OK	10:44
Air Blank	0.000	10:45
Control Test	0.081	10:45
Air Blank	0.000	10:46
Subject Sample # 1	0.138	10:47
Breath Volume	1.566	10:47
Air Blank	0.000	10:47
Air Blank	0.000	10:49
Subject Sample # 2	0.142	10:49
Breath Volume	1.695	10:49
Air Blank	0.000	10:50
Control Test	0.081	10:50
Air Blank	0.000	10:51
Diagnostics Check	OK	10:51

Data Download Date/Time: 04/16/2019 15:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,D,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CAPRIO,Z
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:13

SUBJECT'S LAST NAME: SINGER

DATE/TIME OF TEST: 04/14/2019 18:41:06

Diagnostics Check	OK	18:33
Air Blank	0.000	18:34
Control Test	0.081	18:34
Air Blank	0.000	18:35
Subject Sample # 1	0.257	18:36
Breath Volume	1.988	18:36
Air Blank	0.000	18:36
Air Blank	0.000	18:38
Subject Sample # 2	0.267	18:39
Breath Volume	2.472	18:39
Air Blank	0.000	18:40
Control Test	0.081	18:40
Air Blank	0.000	18:41
Diagnostics Check	OK	18:41

Data Download Date/Time: 04/16/2019 15:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,M,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: EGUIA

DATE/TIME OF TEST: 04/15/2019 01:12:38

Diagnostics Check	OK	01:05
Air Blank	0.000	01:06
Control Test	0.082	01:06
Air Blank	0.000	01:07
Subject Sample # 1	0.122	01:07
Breath Volume	1.316	01:07
Air Blank	0.000	01:08
Air Blank	0.000	01:10
Subject Sample # 2	0.126	01:11
Breath Volume	1.718	01:11
Air Blank	0.000	01:11
Control Test	0.081	01:12
Air Blank	0.000	01:12
Diagnostics Check	OK	01:12

Data Download Date/Time: 04/16/2019 15:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AUGUSTINE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: HENDERSON

DATE/TIME OF TEST: 04/17/2019 04:43:42

Diagnostics Check	OK	04:37
Air Blank	0.000	04:37
Control Test	0.081	04:38
Air Blank	0.000	04:38
Subject Sample # 1	0.157	04:39
Breath Volume	2.953	04:39
Air Blank	0.000	04:39
Air Blank	0.000	04:41
Subject Sample # 2	0.157	04:42
Breath Volume	2.847	04:42
Air Blank	0.000	04:42
Control Test	0.082	04:43
Air Blank	0.000	04:43
Diagnostics Check	OK	04:43

Data Download Date/Time: 05/01/2019 13:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: JAMES

DATE/TIME OF TEST: 04/19/2019 02:18:56

Diagnostics Check	OK	02:12
Air Blank	0.000	02:13
Control Test	0.081	02:13
Air Blank	0.000	02:14
Subject Sample # 1	0.000	02:14
Breath Volume	1.757	02:14
Air Blank	0.000	02:15
Air Blank	0.000	02:17
Subject Sample # 2	0.000	02:17
Breath Volume	2.886	02:17
Air Blank	0.000	02:18
Control Test	0.082	02:18
Air Blank	0.000	02:18
Diagnostics Check	OK	02:18

Data Download Date/Time: 05/01/2019 13:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,D,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 04/24/2019 18:10:17

Diagnostics Check	OK	18:06	
Air Blank	0.000	18:06	
Control Test	0.081	18:07	
Air Blank	0.000	18:07	
Subject Sample	# 1	18:08	Subject Test Refused
Breath Volume	0.000	18:08	
Air Blank	0.000	18:09	
Control Test	0.081	18:09	
Air Blank	0.000	18:10	
Diagnostics Check	OK	18:10	

Data Download Date/Time: 05/01/2019 13:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AUGUSTINE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: RINDERS

DATE/TIME OF TEST: 04/24/2019 23:21:28

Diagnostics Check	OK	23:16	
Air Blank	0.000	23:16	
Control Test	0.081	23:16	
Air Blank	0.000	23:17	
Subject Sample # 1	0.119	23:17	
Breath Volume	3.019	23:17	
Air Blank	0.000	23:18	
Air Blank		23:20	RFI Detect
Air Blank	0.000	23:20	
Control Test	0.082	23:20	
Air Blank	0.000	23:21	
Diagnostics Check	OK	23:21	

Data Download Date/Time: 05/01/2019 13:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AUGUSTINE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: RINDERS

DATE/TIME OF TEST: 04/24/2019 23:29:01

Diagnostics Check	OK	23:22
Air Blank	0.000	23:23
Control Test	0.081	23:23
Air Blank	0.000	23:23
Subject Sample # 1	0.118	23:24
Breath Volume	3.675	23:24
Air Blank	0.000	23:25
Air Blank	0.000	23:26
Subject Sample # 2	0.119	23:27
Breath Volume	3.523	23:27
Air Blank	0.000	23:28
Control Test	0.081	23:28
Air Blank	0.000	23:28
Diagnostics Check	OK	23:29

Data Download Date/Time: 05/01/2019 13:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JENNINGS,Z
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: WALLACE

DATE/TIME OF TEST: 04/28/2019 03:38:49

Diagnostics Check	OK	03:35	
Air Blank	0.000	03:35	
Control Test	0.081	03:35	
Air Blank	0.000	03:36	
Subject Sample	# 1	03:37	Subject Test Refused
Breath Volume	0.078	03:37	
Air Blank	0.000	03:37	
Control Test	0.081	03:38	
Air Blank	0.000	03:38	
Diagnostics Check	OK	03:38	

Data Download Date/Time: 05/01/2019 13:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,RYAN,W
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BROWN,RYAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: PITTMAN

DATE/TIME OF TEST: 04/14/2019 23:51:50

Diagnostics Check	OK	23:45
Air Blank	0.000	23:46
Control Test	0.081	23:46
Air Blank	0.000	23:47
Subject Sample # 1	0.000	23:47
Breath Volume	2.781	23:47
Air Blank	0.000	23:47
Air Blank	0.000	23:49
Subject Sample # 2	0.000	23:50
Breath Volume	2.582	23:50
Air Blank	0.000	23:50
Control Test	0.080	23:51
Air Blank	0.000	23:51
Diagnostics Check	OK	23:51

Data Download Date/Time: 05/05/2019 18:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAGLET, TONY,
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: ROYCE, JAMES
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 03/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: VANDYKE

DATE/TIME OF TEST: 04/11/2019 00:34:51

Diagnostics Check	OK	00:28
Air Blank	0.000	00:29
Control Test	0.079	00:29
Air Blank	0.000	00:29
Subject Sample # 1	0.143	00:30
Breath Volume	1.527	00:30
Air Blank	0.000	00:31
Air Blank	0.000	00:32
Subject Sample # 2	0.144	00:33
Breath Volume	1.511	00:33
Air Blank	0.000	00:33
Control Test	0.080	00:34
Air Blank	0.000	00:34
Diagnostics Check	OK	00:34

Data Download Date/Time: 04/29/2019 16:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAGLE, TONY,
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: STACY, CODY
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 03/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: SOLIS

DATE/TIME OF TEST: 04/12/2019 01:35:35

Diagnostics Check	OK	01:29
Air Blank	0.000	01:29
Control Test	0.079	01:29
Air Blank	0.000	01:30
Subject Sample # 1	0.246	01:31
Breath Volume	3.058	01:31
Air Blank	0.000	01:31
Air Blank	0.000	01:33
Subject Sample # 2	0.228	01:34
Breath Volume	2.640	01:34
Air Blank	0.000	01:34
Control Test	0.080	01:34
Air Blank	0.000	01:35
Diagnostics Check	OK	01:35

Data Download Date/Time: 04/29/2019 16:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,PAUL,A
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: BURGER,ALISHA
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 03/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 12:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:35

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 04/20/2019 13:12:01

Diagnostics Check	OK	13:09	
Air Blank	0.000	13:09	
Control Test	0.080	13:09	
Air Blank	0.000	13:10	
Subject Sample	# 1	13:10	Subject Test Refused
Breath Volume	0.000	13:10	
Air Blank	0.000	13:11	
Control Test	0.080	13:11	
Air Blank	0.000	13:11	
Diagnostics Check	OK	13:12	

Data Download Date/Time: 04/29/2019 16:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAGLE, TONY,
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: STACY, CODY
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 03/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 04/21/2019 01:07:15

Diagnostics Check	OK	01:01
Air Blank	0.000	01:01
Control Test	0.079	01:01
Air Blank	0.000	01:02
Subject Sample # 1	0.082	01:02
Breath Volume	1.492	01:02
Air Blank	0.000	01:03
Air Blank	0.000	01:05
Subject Sample # 2	0.084	01:05
Breath Volume	1.597	01:05
Air Blank	0.000	01:06
Control Test	0.080	01:06
Air Blank	0.000	01:07
Diagnostics Check	OK	01:07

Data Download Date/Time: 04/29/2019 16:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAGLE, TONY,
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: STACY, CODY
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 03/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: WOODFORD

DATE/TIME OF TEST: 04/25/2019 05:09:47

Diagnostics Check	OK	05:06	
Air Blank	0.000	05:07	
Control Test	0.080	05:07	
Air Blank	0.000	05:08	
Subject Sample	# 1	05:08	Subject Test Refused
Breath Volume	0.000	05:08	
Air Blank	0.000	05:08	
Control Test	0.080	05:09	
Air Blank	0.000	05:09	
Diagnostics Check	OK	05:09	

Data Download Date/Time: 04/29/2019 16:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALTER,CRAIG,K
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: TOOMER,SILAS
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 03/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: VANCAMPER

DATE/TIME OF TEST: 04/28/2019 04:26:33

Diagnostics Check	OK	04:19
Air Blank	0.000	04:19
Control Test	0.079	04:20
Air Blank	0.000	04:20
Subject Sample # 1	0.290	04:21
Breath Volume	1.281	04:21
Air Blank	0.000	04:22
Air Blank	0.000	04:23
Subject Sample # 2	0.287	04:24
Breath Volume	2.289	04:24
Air Blank	0.000	04:25
Control Test	0.078	04:25
Air Blank	0.000	04:26
Diagnostics Check	OK	04:26

Data Download Date/Time: 04/29/2019 16:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALTER,CRAIG,K
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: WINLAND,BOBBY
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 03/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 19:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: SANCHEZ VARGAS

DATE/TIME OF TEST: 04/28/2019 20:06:42

Diagnostics Check	OK	20:03	
Air Blank	0.000	20:04	
Control Test	0.080	20:04	
Air Blank	0.000	20:05	
Subject Sample	# 1	20:05	Subject Test Refused
Breath Volume	0.000	20:05	
Air Blank	0.000	20:05	
Control Test	0.079	20:06	
Air Blank	0.000	20:06	
Diagnostics Check	OK	20:06	

Data Download Date/Time: 04/29/2019 16:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHESSHER,SANDRA,W
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: BARGE,SUSAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: CAGLE

DATE/TIME OF TEST: 04/27/2019 23:17:19

Diagnostics Check	OK	23:11
Air Blank	0.000	23:11
Control Test	0.080	23:11
Air Blank	0.000	23:12
Subject Sample # 1	0.164	23:12
Breath Volume	1.343	23:12
Air Blank	0.000	23:13
Air Blank	0.000	23:15
Subject Sample # 2	0.174	23:15
Breath Volume	1.519	23:15
Air Blank	0.000	23:16
Control Test	0.082	23:16
Air Blank	0.000	23:17
Diagnostics Check	OK	23:17

Data Download Date/Time: 05/05/2019 15:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001123

INSTRUMENT REGISTERED TO: FHP TROOP D ORLANDO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRACH,STEFAN,G
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: DRACH,STEFAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:38

SUBJECT'S LAST NAME: PEACOCK

DATE/TIME OF TEST: 04/03/2019 20:14:07

Diagnostics Check	OK	20:08
Air Blank	0.000	20:08
Control Test	0.080	20:09
Air Blank	0.000	20:09
Subject Sample # 1	0.206	20:09
Breath Volume	2.628	20:09
Air Blank	0.000	20:10
Air Blank	0.000	20:12
Subject Sample # 2	0.203	20:12
Breath Volume	2.570	20:12
Air Blank	0.000	20:13
Control Test	0.079	20:13
Air Blank	0.000	20:14
Diagnostics Check	OK	20:14

Data Download Date/Time: 04/19/2019 12:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001123

INSTRUMENT REGISTERED TO: FHP TROOP D ORLANDO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRACH,STEFAN,G
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: DRACH,STEFAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 04/04/2019 02:04:24

Diagnostics Check	OK	01:58
Air Blank	0.000	01:58
Control Test	0.080	01:59
Air Blank	0.000	01:59
Subject Sample # 1	0.169	02:00
Breath Volume	2.847	02:00
Air Blank	0.000	02:00
Air Blank	0.000	02:02
Subject Sample # 2	0.162	02:02
Breath Volume	2.269	02:02
Air Blank	0.000	02:03
Control Test	0.079	02:03
Air Blank	0.000	02:04
Diagnostics Check	OK	02:04

Data Download Date/Time: 04/19/2019 12:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001123

INSTRUMENT REGISTERED TO: FHP TROOP D ORLANDO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRACH,STEFAN,G
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: DRACH,STEFAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: RAMOS HERRERA

DATE/TIME OF TEST: 04/11/2019 03:55:44

Diagnostics Check	OK	03:49
Air Blank	0.000	03:50
Control Test	0.080	03:50
Air Blank	0.000	03:51
Subject Sample # 1	0.054	03:51
Breath Volume	3.175	03:51
Air Blank	0.000	03:52
Air Blank	0.000	03:53
Subject Sample # 2	0.053	03:54
Breath Volume	2.945	03:54
Air Blank	0.000	03:54
Control Test	0.080	03:55
Air Blank	0.000	03:55
Diagnostics Check	OK	03:55

Data Download Date/Time: 04/19/2019 12:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES,STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HAMILTON,ROBERT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: WIEGELE

DATE/TIME OF TEST: 04/04/2019 03:04:49

Diagnostics Check	OK	02:58
Air Blank	0.000	02:59
Control Test	0.080	02:59
Air Blank	0.000	03:00
Subject Sample # 1	0.121	03:00
Breath Volume	2.757	03:00
Air Blank	0.000	03:01
Air Blank	0.000	03:03
Subject Sample # 2	0.121	03:03
Breath Volume	2.636	03:03
Air Blank	0.000	03:03
Control Test	0.080	03:04
Air Blank	0.000	03:04
Diagnostics Check	OK	03:04

Data Download Date/Time: 04/29/2019 17:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES,STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HAMILTON,ROBERT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: COUGHLIN

DATE/TIME OF TEST: 04/04/2019 03:17:53

Diagnostics Check	OK	03:06	
Air Blank	0.000	03:06	
Control Test	0.080	03:07	
Air Blank	0.000	03:07	
Subject Sample	# 1	03:10	No Sample Provided
Breath Volume	0.000	03:10	
Air Blank	0.000	03:11	
Air Blank	0.000	03:13	
Subject Sample	# 2	03:16	No Sample Provided
Breath Volume	0.000	03:16	
Air Blank	0.000	03:17	
Control Test	0.079	03:17	
Air Blank	0.000	03:17	
Diagnostics Check	OK	03:17	

Data Download Date/Time: 04/29/2019 17:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES, STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SYKES, STEVEN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: LETCHWORTH

DATE/TIME OF TEST: 04/14/2019 04:06:56

Diagnostics Check	OK	04:00
Air Blank	0.000	04:00
Control Test	0.080	04:00
Air Blank	0.000	04:01
Subject Sample # 1	0.000	04:02
Breath Volume	1.558	04:02
Air Blank	0.000	04:02
Air Blank	0.000	04:04
Subject Sample # 2	0.000	04:05
Breath Volume	1.359	04:05
Air Blank	0.000	04:06
Control Test	0.081	04:06
Air Blank	0.000	04:06
Diagnostics Check	OK	04:06

Data Download Date/Time: 04/29/2019 17:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHRISTELLO,CLIFFORD,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: CHRISTELLO,CLIFFORD
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: HODGE

DATE/TIME OF TEST: 04/14/2019 05:31:07

Diagnostics Check	OK	05:19	
Air Blank	0.000	05:20	
Control Test	0.080	05:20	
Air Blank	0.000	05:21	
Subject Sample	# 1	05:24	No Sample Provided
Breath Volume	0.000	05:24	
Air Blank	0.000	05:24	
Air Blank	0.000	05:26	
Subject Sample	# 2	05:29	No Sample Provided
Breath Volume	0.000	05:29	
Air Blank	0.000	05:30	
Control Test	0.081	05:30	
Air Blank	0.000	05:31	
Diagnostics Check	OK	05:31	

Data Download Date/Time: 04/29/2019 17:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER, ADAM, H
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: POTTER, ADAM
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: LAHENS

DATE/TIME OF TEST: 04/17/2019 03:49:14

Diagnostics Check	OK	03:46	
Air Blank	0.000	03:46	
Control Test	0.081	03:47	
Air Blank	0.000	03:47	
Subject Sample	# 1	03:47	Subject Test Refused
Breath Volume	0.000	03:47	
Air Blank	0.000	03:48	
Control Test	0.081	03:48	
Air Blank	0.000	03:49	
Diagnostics Check	OK	03:49	

Data Download Date/Time: 04/29/2019 17:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES,STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HIGGEMBOTHAM,JACK
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: CRESPO RODRIGUEZ

DATE/TIME OF TEST: 04/24/2019 04:01:25

Diagnostics Check	OK	03:51
Air Blank	0.000	03:51
Control Test	0.079	03:52
Air Blank	0.000	03:52
Subject Sample # 1	0.142	03:55
Breath Volume	1.828	03:55
Air Blank	0.000	03:56
Air Blank	0.000	03:58
Subject Sample # 2	0.148	03:59
Breath Volume	1.542	03:59
Air Blank	0.000	04:00
Control Test	0.079	04:00
Air Blank	0.000	04:01
Diagnostics Check	OK	04:01

Data Download Date/Time: 04/29/2019 17:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001129

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRENCH, CALEB, L
BREATH TEST OPERATOR AGENCY: Putnam County SO
ARRESTING OFFICER: NUNEZ, MICHAEL
ARRESTING OFFICER AGENCY: Putnam County SO
LAST AGENCY INSPECTION: 04/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A3
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: CROFF

DATE/TIME OF TEST: 04/09/2019 00:48:30

Diagnostics Check	OK	00:41
Air Blank	0.000	00:42
Control Test	0.080	00:42
Air Blank	0.000	00:43
Subject Sample # 1	0.125	00:44
Breath Volume	1.449	00:44
Air Blank	0.000	00:44
Air Blank	0.000	00:46
Subject Sample # 2	0.127	00:47
Breath Volume	1.625	00:47
Air Blank	0.000	00:47
Control Test	0.077	00:47
Air Blank	0.000	00:48
Diagnostics Check	OK	00:48

Data Download Date/Time: 05/12/2019 21:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001129

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,ZACHARY,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: STEPHENS,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A3
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 13:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:16

SUBJECT'S LAST NAME: HORWAT

DATE/TIME OF TEST: 04/19/2019 14:48:44

Diagnostics Check	OK	14:42
Air Blank	0.000	14:42
Control Test	0.079	14:43
Air Blank	0.000	14:43
Subject Sample # 1	0.221	14:44
Breath Volume	2.652	14:44
Air Blank	0.000	14:44
Air Blank	0.000	14:46
Subject Sample # 2	0.201	14:47
Breath Volume	1.113	14:47
Air Blank	0.000	14:47
Control Test	0.077	14:48
Air Blank	0.000	14:48
Diagnostics Check	OK	14:48

Data Download Date/Time: 05/12/2019 21:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001129

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VEGA,KELVIN,
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: BERGMAN,A
ARRESTING OFFICER AGENCY: Palatka PD
LAST AGENCY INSPECTION: 04/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A3
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: STEPHENS

DATE/TIME OF TEST: 04/20/2019 03:35:20

Diagnostics Check	OK	03:26
Air Blank	0.000	03:26
Control Test	0.079	03:26
Air Blank	0.000	03:27
Subject Sample # 1	0.231	03:30
Breath Volume	2.066	03:30
Air Blank	0.000	03:30
Air Blank	0.000	03:32
Subject Sample # 2	0.221	03:33
Breath Volume	1.343	03:33
Air Blank	0.000	03:34
Control Test	0.078	03:34
Air Blank	0.000	03:35
Diagnostics Check	OK	03:35

Data Download Date/Time: 05/12/2019 21:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001129

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS, TIFFANY, R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: WILLIAMS, TIFFANY
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A3
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: LIRA MENDOZA

DATE/TIME OF TEST: 04/25/2019 21:36:16

Diagnostics Check	OK	21:26
Air Blank	0.000	21:26
Control Test	0.080	21:27
Air Blank	0.000	21:27
Subject Sample # 1	0.219	21:30
Breath Volume	2.152	21:30
Air Blank	0.000	21:31
Air Blank	0.000	21:33
Subject Sample # 2	0.221	21:34
Breath Volume	1.253	21:34
Air Blank	0.000	21:35
Control Test	0.078	21:35
Air Blank	0.000	21:36
Diagnostics Check	OK	21:36

Data Download Date/Time: 05/12/2019 21:07