

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

INSTRUMENT REGISTERED TO: HIALEAH GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUERTA, EDWIN,
BREATH TEST OPERATOR AGENCY: HIALEAH GARDENS PD
ARRESTING OFFICER: HUERTA, EDWIN
ARRESTING OFFICER AGENCY: HIALEAH GARDENS PD
LAST AGENCY INSPECTION: 05/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: VIVES

DATE/TIME OF TEST: 06/22/2019 00:10:26

Diagnostics Check		OK	00:02	
Air Blank		0.000	00:02	
Control Test		0.078	00:02	
Air Blank		0.000	00:03	
Subject Sample	# 1	0.235	00:06	Volume Not Met
Breath Volume		0.804	00:06	
Air Blank		0.000	00:07	
Air Blank		0.000	00:08	
Subject Sample	# 2		00:09	Subject Test Refused
Breath Volume		0.000	00:09	
Air Blank		0.000	00:09	
Control Test		0.076	00:09	
Air Blank		0.000	00:10	
Diagnostics Check		OK	00:10	

Data Download Date/Time: 06/25/2019 14:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTIERREZ,KRISTINA,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: GUTIERREZ,KRISTINA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: ORTA

DATE/TIME OF TEST: 06/04/2019 00:56:08

Diagnostics Check	OK	00:50	
Air Blank	0.000	00:51	
Control Test	0.077	00:51	
Air Blank	0.000	00:51	
Subject Sample	# 1	00:54	Subject Test Refused
Breath Volume	0.000	00:54	
Air Blank	0.000	00:55	
Control Test	0.078	00:55	
Air Blank	0.000	00:56	
Diagnostics Check	OK	00:56	

Data Download Date/Time: 06/28/2019 10:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,D,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: VARGAS,D
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 06/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 06/14/2019 23:56:16

Diagnostics Check	OK	23:49
Air Blank	0.000	23:50
Control Test	0.079	23:50
Air Blank	0.000	23:51
Subject Sample # 1	0.165	23:51
Breath Volume	1.160	23:51
Air Blank	0.000	23:52
Air Blank	0.000	23:54
Subject Sample # 2	0.171	23:54
Breath Volume	2.210	23:54
Air Blank	0.000	23:55
Control Test	0.080	23:55
Air Blank	0.000	23:56
Diagnostics Check	OK	23:56

Data Download Date/Time: 07/11/2019 09:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NADAL,JOSE,V
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: NADAL,JOSE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 06/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 06/15/2019 04:48:28

Diagnostics Check	OK	04:33	
Air Blank	0.000	04:33	
Control Test	0.079	04:33	
Air Blank	0.000	04:34	
Subject Sample # 1	0.162	04:36	
Breath Volume	1.425	04:36	
Air Blank	0.000	04:37	
Air Blank	0.000	04:38	
Subject Sample # 2	0.132	04:41	Volume Not Met
Breath Volume	0.687	04:41	
Air Blank	0.000	04:43	
Air Blank	0.000	04:44	
Subject Sample # 3	0.152	04:46	
Breath Volume	1.277	04:46	
Air Blank	0.000	04:47	
Control Test	0.073	04:47	Control Outside Tolerance
Air Blank	0.000	04:48	

Data Download Date/Time: 07/11/2019 09:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NADAL,JOSE,V
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: NADAL,JOSE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 06/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 06/15/2019 05:02:00

Diagnostics Check	OK	04:51
Air Blank	0.000	04:52
Control Test	0.077	04:52
Air Blank	0.000	04:53
Subject Sample # 1	0.151	04:54
Breath Volume	1.273	04:54
Air Blank	0.000	04:55
Air Blank	0.000	04:57
Subject Sample # 2	0.155	05:00
Breath Volume	1.937	05:00
Air Blank	0.000	05:01
Control Test	0.079	05:01
Air Blank	0.000	05:01
Diagnostics Check	OK	05:02

Data Download Date/Time: 07/11/2019 09:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POVEDA,EMILIE,
BREATH TEST OPERATOR AGENCY: SWEETWATER PD
ARRESTING OFFICER: FERNANDEZ,ORLANDO
ARRESTING OFFICER AGENCY: SWEETWATER PD
LAST AGENCY INSPECTION: 06/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: TREJOS CASAYA

DATE/TIME OF TEST: 06/16/2019 03:44:06

Diagnostics Check	OK	03:37
Air Blank	0.000	03:37
Control Test	0.080	03:38
Air Blank	0.000	03:38
Subject Sample # 1	0.088	03:39
Breath Volume	1.812	03:39
Air Blank	0.000	03:40
Air Blank	0.000	03:42
Subject Sample # 2	0.086	03:42
Breath Volume	1.882	03:42
Air Blank	0.000	03:43
Control Test	0.078	03:43
Air Blank	0.000	03:44
Diagnostics Check	OK	03:44

Data Download Date/Time: 07/11/2019 09:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTIERREZ,KRISTINA,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: GUTIERREZ,KRISTINA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 06/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: USEDA ARAICA

DATE/TIME OF TEST: 06/16/2019 04:05:11

Diagnostics Check	OK	04:00	
Air Blank	0.000	04:01	
Control Test	0.079	04:01	
Air Blank	0.000	04:01	
Subject Sample	# 1	04:03	Subject Test Refused
Breath Volume	0.000	04:03	
Air Blank	0.000	04:04	
Control Test	0.079	04:04	
Air Blank	0.000	04:05	
Diagnostics Check	OK	04:05	

Data Download Date/Time: 07/11/2019 09:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NADAL,JOSE,V
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: SUAREZ,CARLOS
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 06/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: GUILLEN

DATE/TIME OF TEST: 06/16/2019 04:15:21

Diagnostics Check	OK	04:12	
Air Blank	0.000	04:12	
Control Test	0.079	04:13	
Air Blank	0.000	04:13	
Subject Sample	# 1	04:13	Subject Test Refused
Breath Volume	0.000	04:13	
Air Blank	0.000	04:14	
Control Test	0.080	04:14	
Air Blank	0.000	04:15	
Diagnostics Check	OK	04:15	

Data Download Date/Time: 07/11/2019 09:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTANA,JORGE,F
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: QUINTANA,JORGE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 06/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: GONZALEZ-MAZONIS

DATE/TIME OF TEST: 06/16/2019 04:50:46

Diagnostics Check		OK	04:39	
Air Blank		0.000	04:39	
Control Test		0.080	04:39	
Air Blank		0.000	04:40	
Subject Sample	# 1	0.106	04:43	Volume Not Met
Breath Volume		0.648	04:43	
Air Blank		0.000	04:44	
Air Blank		0.000	04:45	
Subject Sample	# 2	0.085	04:49	Volume Not Met
Breath Volume		0.050	04:49	
Air Blank		0.000	04:49	
Control Test		0.079	04:50	
Air Blank		0.000	04:50	
Diagnostics Check		OK	04:50	

Data Download Date/Time: 07/11/2019 09:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTANA,JORGE,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: QUINTANA,JORGE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 06/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: GONZALEZ-MAZONIS

DATE/TIME OF TEST: 06/16/2019 04:55:25

Diagnostics Check	OK	04:52	
Air Blank	0.000	04:53	
Control Test	0.078	04:53	
Air Blank	0.000	04:53	
Subject Sample	# 1	04:54	Subject Test Refused
Breath Volume	0.000	04:54	
Air Blank	0.000	04:54	
Control Test	0.078	04:54	
Air Blank	0.000	04:55	
Diagnostics Check	OK	04:55	

Data Download Date/Time: 07/11/2019 09:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 06/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 05:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:53

SUBJECT'S LAST NAME: GEORGE MAJEROWICZ

DATE/TIME OF TEST: 06/19/2019 07:24:36

Diagnostics Check	OK	07:18
Air Blank	0.000	07:18
Control Test	0.079	07:19
Air Blank	0.000	07:19
Subject Sample # 1	0.031	07:20
Breath Volume	1.960	07:20
Air Blank	0.000	07:20
Air Blank	0.000	07:22
Subject Sample # 2	0.031	07:23
Breath Volume	2.406	07:23
Air Blank	0.000	07:23
Control Test	0.080	07:23
Air Blank	0.000	07:24
Diagnostics Check	OK	07:24

Data Download Date/Time: 07/11/2019 09:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: CLOSIUS,R
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 06/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 05:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:15

SUBJECT'S LAST NAME: BETHEL

DATE/TIME OF TEST: 06/20/2019 06:47:52

Diagnostics Check	OK	06:41
Air Blank	0.000	06:42
Control Test	0.079	06:42
Air Blank	0.000	06:42
Subject Sample # 1	0.000	06:43
Breath Volume	2.406	06:43
Air Blank	0.000	06:44
Air Blank	0.000	06:46
Subject Sample # 2	0.000	06:46
Breath Volume	2.125	06:46
Air Blank	0.000	06:46
Control Test	0.080	06:47
Air Blank	0.000	06:47
Diagnostics Check	OK	06:47

Data Download Date/Time: 07/11/2019 09:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARDELAS,MAYLE,
BREATH TEST OPERATOR AGENCY: OTHER LOCAL
ARRESTING OFFICER: SIDDO,RAAJU
ARRESTING OFFICER AGENCY: OTHER LOCAL
LAST AGENCY INSPECTION: 06/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: SANCHEZ-VARGAS

DATE/TIME OF TEST: 06/23/2019 05:29:20

Diagnostics Check	OK	05:25	
Air Blank	0.000	05:25	
Control Test	0.080	05:25	
Air Blank	0.000	05:26	
Subject Sample	# 1	05:27	Subject Test Refused
Breath Volume	0.000	05:27	
Air Blank	0.000	05:28	
Control Test	0.079	05:28	
Air Blank	0.000	05:29	
Diagnostics Check	OK	05:29	

Data Download Date/Time: 07/11/2019 09:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANCO,GREGORY,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 06/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 05:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:07

SUBJECT'S LAST NAME: MENA-MURILLO

DATE/TIME OF TEST: 06/23/2019 06:48:17

Diagnostics Check		OK	06:40
Air Blank		0.000	06:40
Control Test		0.080	06:40
Air Blank		0.000	06:41
Subject Sample	# 1	0.201	06:42
Breath Volume		2.031	06:42
Air Blank		0.000	06:43
Air Blank		0.000	06:45
Subject Sample	# 2	0.190	06:46
Breath Volume		1.714	06:46
Air Blank		0.000	06:47
Control Test		0.079	06:47
Air Blank		0.000	06:48
Diagnostics Check		OK	06:48

Data Download Date/Time: 07/11/2019 09:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAINTILIEEN,BOB,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: SAINTILIEEN,BOB
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 06/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: RODRIGUEZ GUZMAN

DATE/TIME OF TEST: 06/23/2019 20:21:26

Diagnostics Check	OK	20:18	
Air Blank	0.000	20:18	
Control Test	0.079	20:19	
Air Blank	0.000	20:19	
Subject Sample	# 1	20:20	Subject Test Refused
Breath Volume	0.000	20:20	
Air Blank	0.000	20:20	
Control Test	0.079	20:20	
Air Blank	0.000	20:21	
Diagnostics Check	OK	20:21	

Data Download Date/Time: 07/11/2019 09:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTIERREZ,KRISTINA,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: GUTIERREZ,KRISTINA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 06/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 06/24/2019 03:21:59

Diagnostics Check	OK	03:17	
Air Blank	0.000	03:17	
Control Test	0.079	03:18	
Air Blank	0.000	03:18	
Subject Sample	# 1	03:20	Subject Test Refused
Breath Volume	0.000	03:20	
Air Blank	0.000	03:21	
Control Test	0.079	03:21	
Air Blank	0.000	03:21	
Diagnostics Check	OK	03:21	

Data Download Date/Time: 07/11/2019 09:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUJILLO,GEORGE,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: TRUJILLO,GEORGE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 06/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:19

SUBJECT'S LAST NAME: OCAMPO

DATE/TIME OF TEST: 06/25/2019 17:52:41

Diagnostics Check		OK	17:39	
Air Blank		0.000	17:40	
Control Test		0.079	17:40	
Air Blank		0.000	17:41	
Subject Sample	# 1	0.269	17:44	Volume Not Met
Breath Volume		0.019	17:44	
Air Blank		0.000	17:45	
Air Blank		0.000	17:46	
Subject Sample	# 2	0.259	17:47	
Breath Volume		2.027	17:47	
Air Blank		0.000	17:48	
Air Blank		0.000	17:50	
Subject Sample	# 3	0.274	17:50	
Breath Volume		1.484	17:50	
Air Blank		0.000	17:51	
Control Test		0.075	17:52	
Air Blank		0.000	17:52	
Diagnostics Check		OK	17:52	

Data Download Date/Time: 07/11/2019 09:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTIERREZ,KRISTINA,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: GUTIERREZ,KRISTINA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 06/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: PADILLA

DATE/TIME OF TEST: 06/26/2019 00:44:45

Diagnostics Check	OK	00:41	
Air Blank	0.000	00:41	
Control Test	0.079	00:42	
Air Blank	0.000	00:42	
Subject Sample	# 1	00:43	Subject Test Refused
Breath Volume	0.000	00:43	
Air Blank	0.000	00:43	
Control Test	0.080	00:44	
Air Blank	0.000	00:44	
Diagnostics Check	OK	00:44	

Data Download Date/Time: 07/11/2019 09:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POVEDA,EMILIE,
BREATH TEST OPERATOR AGENCY: SWEETWATER PD
ARRESTING OFFICER: POVEDA,EMILIE
ARRESTING OFFICER AGENCY: SWEETWATER PD
LAST AGENCY INSPECTION: 06/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 06/26/2019 03:40:16

Diagnostics Check	OK	03:30
Air Blank	0.000	03:31
Control Test	0.079	03:31
Air Blank	0.000	03:32
Subject Sample # 1	0.061	03:33
Breath Volume	1.500	03:33
Air Blank	0.000	03:34
Air Blank	0.000	03:36
Subject Sample # 2	0.062	03:38
Breath Volume	3.082	03:38
Air Blank	0.000	03:39
Control Test	0.077	03:39
Air Blank	0.000	03:40
Diagnostics Check	OK	03:40

Data Download Date/Time: 07/11/2019 09:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORENZO, YOSNIEL,
BREATH TEST OPERATOR AGENCY: SWEETWATER PD
ARRESTING OFFICER: LORENZO, YOSNIEL
ARRESTING OFFICER AGENCY: SWEETWATER PD
LAST AGENCY INSPECTION: 06/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: GUTIERREZ

DATE/TIME OF TEST: 06/27/2019 18:47:27

Diagnostics Check	OK	18:40
Air Blank	0.000	18:41
Control Test	0.080	18:41
Air Blank	0.000	18:41
Subject Sample # 1	0.288	18:42
Breath Volume	1.453	18:42
Air Blank	0.000	18:43
Air Blank	0.000	18:44
Subject Sample # 2	0.304	18:45
Breath Volume	2.031	18:45
Air Blank	0.000	18:46
Control Test	0.078	18:46
Air Blank	0.000	18:47
Diagnostics Check	OK	18:47

Data Download Date/Time: 07/11/2019 09:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERRER,RAFAEL,A
BREATH TEST OPERATOR AGENCY: OTHER LOCAL
ARRESTING OFFICER: FERRER,RAFAEL
ARRESTING OFFICER AGENCY: OTHER LOCAL
LAST AGENCY INSPECTION: 06/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 05:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:18

SUBJECT'S LAST NAME: LOLA

DATE/TIME OF TEST: 06/28/2019 05:50:03

Diagnostics Check	OK	05:39	
Air Blank	0.000	05:39	
Control Test	0.079	05:40	
Air Blank	0.000	05:40	
Subject Sample	# 1	05:43	No Sample Provided
Breath Volume	0.000	05:43	
Air Blank	0.000	05:44	
Air Blank	0.000	05:46	
Subject Sample	# 2	05:48	Subject Test Refused
Breath Volume	0.000	05:48	
Air Blank	0.000	05:49	
Control Test	0.080	05:49	
Air Blank	0.000	05:49	
Diagnostics Check	OK	05:50	

Data Download Date/Time: 07/11/2019 09:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,D,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: VARGAS,D
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 06/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: STUBBS

DATE/TIME OF TEST: 06/28/2019 23:09:30

Diagnostics Check		OK	23:01
Air Blank		0.000	23:01
Control Test		0.080	23:02
Air Blank		0.000	23:02
Subject Sample	# 1	0.081	23:04
Breath Volume		1.296	23:04
Air Blank		0.000	23:05
Air Blank		0.000	23:07
Subject Sample	# 2	0.094	23:07
Breath Volume		2.089	23:07
Air Blank		0.000	23:08
Control Test		0.079	23:08
Air Blank		0.000	23:09
Diagnostics Check		OK	23:09

Data Download Date/Time: 07/11/2019 09:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTIERREZ,KRISTINA,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: GUTIERREZ,KRISTINA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 06/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: DEL ALAMO BACHS

DATE/TIME OF TEST: 06/29/2019 02:13:26

Diagnostics Check	OK	02:06
Air Blank	0.000	02:06
Control Test	0.080	02:06
Air Blank	0.000	02:07
Subject Sample # 1	0.210	02:08
Breath Volume	2.515	02:08
Air Blank	0.000	02:09
Air Blank	0.000	02:10
Subject Sample # 2	0.217	02:11
Breath Volume	2.539	02:11
Air Blank	0.000	02:12
Control Test	0.079	02:12
Air Blank	0.000	02:13
Diagnostics Check	OK	02:13

Data Download Date/Time: 07/11/2019 09:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIDDOR,RAAJU,S
BREATH TEST OPERATOR AGENCY: OTHER LOCAL
ARRESTING OFFICER: SIDDO,RAAJU
ARRESTING OFFICER AGENCY: OTHER LOCAL
LAST AGENCY INSPECTION: 06/13/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: TRUCHSESS RUBIO

DATE/TIME OF TEST: 06/30/2019 01:09:54

Diagnostics Check	OK	01:02
Air Blank	0.000	01:03
Control Test	0.080	01:03
Air Blank	0.000	01:03
Subject Sample # 1	0.167	01:04
Breath Volume	1.921	01:04
Air Blank	0.000	01:05
Air Blank	0.000	01:07
Subject Sample # 2	0.155	01:08
Breath Volume	1.394	01:08
Air Blank	0.000	01:08
Control Test	0.079	01:09
Air Blank	0.000	01:09
Diagnostics Check	OK	01:09

Data Download Date/Time: 07/11/2019 09:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTIERREZ,KRISTINA,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: GUTIERREZ,KRISTINA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 06/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: KRINGS

DATE/TIME OF TEST: 06/30/2019 03:45:24

Diagnostics Check	OK	03:30	
Air Blank	0.000	03:31	
Control Test	0.080	03:31	
Air Blank	0.000	03:32	
Subject Sample	# 1	03:35	No Sample Provided
Breath Volume	0.000	03:35	
Air Blank	0.000	03:35	
Air Blank	0.000	03:37	
Subject Sample	# 2	0.171	03:40
Breath Volume	1.781	03:40	
Air Blank	0.000	03:41	
Air Blank	0.000	03:42	
Subject Sample	# 3	0.168	03:43
Breath Volume	1.351	03:43	
Air Blank	0.000	03:44	
Control Test	0.077	03:44	
Air Blank	0.000	03:45	
Diagnostics Check	OK	03:45	

Data Download Date/Time: 07/11/2019 09:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARDELAS,MAYLE,
BREATH TEST OPERATOR AGENCY: OTHER LOCAL
ARRESTING OFFICER: BARDELAS,MAYLE
ARRESTING OFFICER AGENCY: OTHER LOCAL
LAST AGENCY INSPECTION: 06/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: MORALES CRUZ

DATE/TIME OF TEST: 06/30/2019 05:26:54

Diagnostics Check		OK	05:20
Air Blank		0.000	05:20
Control Test		0.079	05:20
Air Blank		0.000	05:21
Subject Sample	# 1	0.206	05:22
Breath Volume		2.597	05:22
Air Blank		0.000	05:23
Air Blank		0.000	05:24
Subject Sample	# 2	0.197	05:25
Breath Volume		1.894	05:25
Air Blank		0.000	05:25
Control Test		0.075	05:26
Air Blank		0.000	05:26
Diagnostics Check		OK	05:26

Data Download Date/Time: 07/11/2019 09:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIDDO,RAAJU,S
BREATH TEST OPERATOR AGENCY: OTHER LOCAL
ARRESTING OFFICER: ZORRILLA,JYNX
ARRESTING OFFICER AGENCY: OTHER LOCAL
LAST AGENCY INSPECTION: 06/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 08:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:36

SUBJECT'S LAST NAME: BONILLA DIAZ

DATE/TIME OF TEST: 06/30/2019 08:59:55

Diagnostics Check	OK	08:57	
Air Blank	0.000	08:57	
Control Test	0.079	08:57	
Air Blank	0.000	08:58	
Subject Sample	# 1	08:58	Subject Test Refused
Breath Volume	0.000	08:58	
Air Blank	0.000	08:58	
Control Test	0.079	08:59	
Air Blank	0.000	08:59	
Diagnostics Check	OK	08:59	

Data Download Date/Time: 07/11/2019 09:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARIDA,K,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EDWARDS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 06:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:12

SUBJECT'S LAST NAME: ARMSTRONG

DATE/TIME OF TEST: 06/01/2019 08:39:51

Diagnostics Check	OK	08:33
Air Blank	0.000	08:33
Control Test	0.081	08:34
Air Blank	0.000	08:34
Subject Sample # 1	0.031	08:35
Breath Volume	1.781	08:35
Air Blank	0.000	08:36
Air Blank	0.000	08:37
Subject Sample # 2	0.033	08:38
Breath Volume	1.839	08:38
Air Blank	0.000	08:38
Control Test	0.081	08:39
Air Blank	0.000	08:39
Diagnostics Check	OK	08:39

Data Download Date/Time: 06/12/2019 12:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARIDA,K,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISHER,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 08:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:22

SUBJECT'S LAST NAME: DANIELS

DATE/TIME OF TEST: 06/02/2019 10:48:47

Diagnostics Check	OK	10:42
Air Blank	0.000	10:43
Control Test	0.081	10:43
Air Blank	0.000	10:43
Subject Sample # 1	0.140	10:44
Breath Volume	3.070	10:44
Air Blank	0.000	10:44
Air Blank	0.000	10:46
Subject Sample # 2	0.133	10:47
Breath Volume	3.023	10:47
Air Blank	0.000	10:47
Control Test	0.081	10:48
Air Blank	0.000	10:48
Diagnostics Check	OK	10:48

Data Download Date/Time: 06/12/2019 12:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: NGUYEN

DATE/TIME OF TEST: 06/13/2019 01:08:20

Diagnostics Check	OK	01:02
Air Blank	0.000	01:02
Control Test	0.080	01:02
Air Blank	0.000	01:03
Subject Sample # 1	0.068	01:03
Breath Volume	3.066	01:03
Air Blank	0.000	01:04
Air Blank	0.000	01:06
Subject Sample # 2	0.070	01:06
Breath Volume	2.664	01:06
Air Blank	0.000	01:07
Control Test	0.081	01:07
Air Blank	0.000	01:08
Diagnostics Check	OK	01:08

Data Download Date/Time: 06/26/2019 13:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: CAMERON

DATE/TIME OF TEST: 06/13/2019 02:15:17

Diagnostics Check	OK	02:09
Air Blank	0.000	02:09
Control Test	0.081	02:09
Air Blank	0.000	02:10
Subject Sample # 1	0.000	02:10
Breath Volume	2.984	02:10
Air Blank	0.000	02:11
Air Blank	0.000	02:13
Subject Sample # 2	0.000	02:13
Breath Volume	2.226	02:13
Air Blank	0.000	02:14
Control Test	0.080	02:14
Air Blank	0.000	02:15
Diagnostics Check	OK	02:15

Data Download Date/Time: 06/26/2019 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: SOLES

DATE/TIME OF TEST: 06/15/2019 19:35:00

Diagnostics Check	OK	19:31	
Air Blank	0.000	19:31	
Control Test	0.080	19:31	
Air Blank	0.000	19:32	
Subject Sample	# 1	19:33	Subject Test Refused
Breath Volume	0.000	19:33	
Air Blank	0.000	19:34	
Control Test	0.080	19:34	
Air Blank	0.000	19:34	
Diagnostics Check	OK	19:35	

Data Download Date/Time: 06/26/2019 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GHENT,STANLEY,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: GHENT

DATE/TIME OF TEST: 06/17/2019 02:59:35

Diagnostics Check	OK	02:55	
Air Blank	0.000	02:56	
Control Test	0.080	02:56	
Air Blank	0.000	02:56	
Subject Sample	# 1	02:58	Subject Test Refused
Breath Volume	0.000	02:58	
Air Blank	0.000	02:58	
Control Test	0.080	02:58	
Air Blank	0.000	02:59	
Diagnostics Check	OK	02:59	

Data Download Date/Time: 06/26/2019 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 06/01/2019 04:13:49

Diagnostics Check	OK	04:11	
Air Blank	0.000	04:11	
Control Test	0.080	04:11	
Air Blank	0.000	04:12	
Subject Sample	# 1	04:12	Subject Test Refused
Breath Volume	0.000	04:12	
Air Blank	0.000	04:12	
Control Test	0.080	04:13	
Air Blank	0.000	04:13	
Diagnostics Check	OK	04:13	

Data Download Date/Time: 06/12/2019 11:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:26

SUBJECT'S LAST NAME: BATES

DATE/TIME OF TEST: 06/01/2019 20:49:37

Diagnostics Check	OK	20:46	
Air Blank	0.000	20:47	
Control Test	0.079	20:47	
Air Blank	0.000	20:47	
Subject Sample	# 1	20:48	Subject Test Refused
Breath Volume	0.000	20:48	
Air Blank	0.000	20:48	
Control Test	0.079	20:49	
Air Blank	0.000	20:49	
Diagnostics Check	OK	20:49	

Data Download Date/Time: 06/12/2019 11:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: GARCIA LOVATO

DATE/TIME OF TEST: 06/02/2019 02:48:15

Diagnostics Check	OK	02:38
Air Blank	0.000	02:38
Control Test	0.080	02:39
Air Blank	0.000	02:39
Subject Sample # 1	0.190	02:40
Breath Volume	1.359	02:40
Air Blank	0.000	02:40
Air Blank	0.000	02:42
Subject Sample # 2	0.159	02:43
Breath Volume	1.113	02:43
Air Blank	0.000	02:44
Air Blank	0.000	02:46
Subject Sample # 3	0.176	02:46
Breath Volume	1.386	02:46
Air Blank	0.000	02:47
Control Test	0.078	02:47
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

Data Download Date/Time: 06/12/2019 11:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FILBERT,E
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 05/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:18

SUBJECT'S LAST NAME: SUNKIREDDY

DATE/TIME OF TEST: 06/02/2019 05:45:41

Diagnostics Check	OK	05:42	
Air Blank	0.000	05:43	
Control Test	0.080	05:43	
Air Blank	0.000	05:44	
Subject Sample	# 1	05:44	Subject Test Refused
Breath Volume	0.000	05:44	
Air Blank	0.000	05:44	
Control Test	0.079	05:45	
Air Blank	0.000	05:45	
Diagnostics Check	OK	05:45	

Data Download Date/Time: 06/12/2019 11:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: WAMPLER

DATE/TIME OF TEST: 06/03/2019 00:02:22

Diagnostics Check		OK	23:50	
Air Blank		0.000	23:51	
Control Test		0.080	23:51	
Air Blank		0.000	23:52	
Subject Sample	# 2	0.105	23:55	Volume Not Met
Breath Volume		0.042	23:55	
Air Blank		0.000	23:55	
Air Blank		0.000	23:57	
Subject Sample	# 1	0.085	00:00	Volume Not Met
Breath Volume		0.312	00:00	
Air Blank		0.000	00:01	
Control Test		0.079	00:01	
Air Blank		0.000	00:02	
Diagnostics Check		OK	00:02	

Data Download Date/Time: 06/12/2019 11:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MCEWAN,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: SUTTON

DATE/TIME OF TEST: 06/03/2019 03:02:30

Diagnostics Check	OK	02:56
Air Blank	0.000	02:56
Control Test	0.079	02:57
Air Blank	0.000	02:57
Subject Sample # 1	0.141	02:58
Breath Volume	1.980	02:58
Air Blank	0.000	02:58
Air Blank	0.000	03:00
Subject Sample # 2	0.135	03:00
Breath Volume	1.207	03:00
Air Blank	0.000	03:01
Control Test	0.079	03:01
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

Data Download Date/Time: 06/12/2019 11:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PARMENTER,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: MYERS

DATE/TIME OF TEST: 06/03/2019 04:15:35

Diagnostics Check	OK	04:09
Air Blank	0.000	04:09
Control Test	0.079	04:10
Air Blank	0.000	04:10
Subject Sample # 1	0.121	04:11
Breath Volume	2.359	04:11
Air Blank	0.000	04:11
Air Blank	0.000	04:13
Subject Sample # 2	0.115	04:14
Breath Volume	2.488	04:14
Air Blank	0.000	04:14
Control Test	0.079	04:14
Air Blank	0.000	04:15
Diagnostics Check	OK	04:15

Data Download Date/Time: 06/12/2019 11:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SOSA,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:50

SUBJECT'S LAST NAME: CLAROS CHINCHILLA

DATE/TIME OF TEST: 06/03/2019 06:17:02

Diagnostics Check	OK	06:10
Air Blank	0.000	06:11
Control Test	0.080	06:11
Air Blank	0.000	06:12
Subject Sample # 1	0.174	06:12
Breath Volume	3.296	06:12
Air Blank	0.000	06:13
Air Blank	0.000	06:15
Subject Sample # 2	0.167	06:15
Breath Volume	2.222	06:15
Air Blank	0.000	06:16
Control Test	0.079	06:16
Air Blank	0.000	06:16
Diagnostics Check	OK	06:17

Data Download Date/Time: 06/12/2019 11:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FILBERT,E
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 05/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: DORE

DATE/TIME OF TEST: 06/06/2019 02:28:43

Diagnostics Check		OK	02:22
Air Blank		0.000	02:23
Control Test		0.080	02:23
Air Blank		0.000	02:23
Subject Sample	# 1	0.000	02:24
Breath Volume		4.496	02:24
Air Blank		0.000	02:24
Air Blank		0.000	02:26
Subject Sample	# 2	0.000	02:27
Breath Volume		4.152	02:27
Air Blank		0.000	02:27
Control Test		0.079	02:28
Air Blank		0.000	02:28
Diagnostics Check		OK	02:28

Data Download Date/Time: 06/12/2019 11:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: ESCOBAR BAUTISTA

DATE/TIME OF TEST: 06/06/2019 23:20:11

Diagnostics Check	OK	23:17	
Air Blank	0.000	23:17	
Control Test	0.080	23:18	
Air Blank	0.000	23:18	
Subject Sample	# 1	23:18	Subject Test Refused
Breath Volume	0.000	23:18	
Air Blank	0.000	23:19	
Control Test	0.080	23:19	
Air Blank	0.000	23:20	
Diagnostics Check	OK	23:20	

Data Download Date/Time: 06/12/2019 12:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: PATTERSON

DATE/TIME OF TEST: 06/08/2019 03:16:05

Diagnostics Check	OK	03:09
Air Blank	0.000	03:10
Control Test	0.080	03:10
Air Blank	0.000	03:11
Subject Sample # 1	0.135	03:11
Breath Volume	1.761	03:11
Air Blank	0.000	03:12
Air Blank	0.000	03:14
Subject Sample # 2	0.140	03:14
Breath Volume	2.093	03:14
Air Blank	0.000	03:15
Control Test	0.079	03:15
Air Blank	0.000	03:16
Diagnostics Check	OK	03:16

Data Download Date/Time: 06/12/2019 12:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: GASKINS

DATE/TIME OF TEST: 06/09/2019 01:37:51

Diagnostics Check	OK	01:35	
Air Blank	0.000	01:35	
Control Test	0.080	01:35	
Air Blank	0.000	01:36	
Subject Sample	# 1	01:36	Subject Test Refused
Breath Volume	0.000	01:36	
Air Blank	0.000	01:36	
Control Test	0.080	01:37	
Air Blank	0.000	01:37	
Diagnostics Check	OK	01:37	

Data Download Date/Time: 06/12/2019 12:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNSON,G
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: PIERCE

DATE/TIME OF TEST: 06/10/2019 00:01:57

Diagnostics Check	OK	23:59	
Air Blank	0.000	23:59	
Control Test	0.079	23:59	
Air Blank	0.000	00:00	
Subject Sample	# 1	00:00	Subject Test Refused
Breath Volume	0.000	00:00	
Air Blank	0.000	00:01	
Control Test	0.079	00:01	
Air Blank	0.000	00:01	
Diagnostics Check	OK	00:01	

Data Download Date/Time: 06/12/2019 12:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: TISDALE

DATE/TIME OF TEST: 06/10/2019 04:42:53

Diagnostics Check	OK	04:36
Air Blank	0.000	04:37
Control Test	0.080	04:37
Air Blank	0.000	04:38
Subject Sample # 1	0.117	04:38
Breath Volume	4.453	04:38
Air Blank	0.000	04:39
Air Blank	0.000	04:40
Subject Sample # 2	0.118	04:41
Breath Volume	4.292	04:41
Air Blank	0.000	04:41
Control Test	0.079	04:42
Air Blank	0.000	04:42
Diagnostics Check	OK	04:42

Data Download Date/Time: 06/12/2019 12:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: LOPEZ-VASQUEZ

DATE/TIME OF TEST: 06/12/2019 02:15:52

Diagnostics Check	OK	02:07
Air Blank	0.000	02:07
Control Test	0.079	02:08
Air Blank	0.000	02:08
Subject Sample # 1	0.147	02:11
Breath Volume	1.214	02:11
Air Blank	0.000	02:11
Air Blank	0.000	02:13
Subject Sample # 2	0.148	02:14
Breath Volume	1.511	02:14
Air Blank	0.000	02:14
Control Test	0.077	02:15
Air Blank	0.000	02:15
Diagnostics Check	OK	02:15

Data Download Date/Time: 06/12/2019 12:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: LANGENBAHN

DATE/TIME OF TEST: 06/12/2019 23:20:41

Diagnostics Check	OK	23:17	
Air Blank	0.000	23:18	
Control Test	0.079	23:18	
Air Blank	0.000	23:19	
Subject Sample	# 1	23:19	Subject Test Refused
Breath Volume	0.000	23:19	
Air Blank	0.000	23:19	
Control Test	0.079	23:20	
Air Blank	0.000	23:20	
Diagnostics Check	OK	23:20	

Data Download Date/Time: 06/26/2019 13:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: SAGGESE

DATE/TIME OF TEST: 06/15/2019 01:07:20

Diagnostics Check	OK	01:04	
Air Blank	0.000	01:04	
Control Test	0.080	01:05	
Air Blank	0.000	01:05	
Subject Sample	# 1	01:05	Subject Test Refused
Breath Volume	0.000	01:05	
Air Blank	0.000	01:06	
Control Test	0.080	01:06	
Air Blank	0.000	01:07	
Diagnostics Check	OK	01:07	

Data Download Date/Time: 06/26/2019 13:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CAMACHO,K
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 06/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: GREENE

DATE/TIME OF TEST: 06/15/2019 03:29:18

Diagnostics Check	OK	03:26	
Air Blank	0.000	03:26	
Control Test	0.079	03:27	
Air Blank	0.000	03:27	
Subject Sample	# 1	03:27	Subject Test Refused
Breath Volume	0.000	03:27	
Air Blank	0.000	03:28	
Control Test	0.080	03:28	
Air Blank	0.000	03:29	
Diagnostics Check	OK	03:29	

Data Download Date/Time: 06/26/2019 13:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: PEREZ-ORTIZ

DATE/TIME OF TEST: 06/18/2019 02:22:36

Diagnostics Check	OK	02:19	
Air Blank	0.000	02:20	
Control Test	0.079	02:20	
Air Blank	0.000	02:21	
Subject Sample	# 1	02:21	Subject Test Refused
Breath Volume	0.000	02:21	
Air Blank	0.000	02:21	
Control Test	0.080	02:22	
Air Blank	0.000	02:22	
Diagnostics Check	OK	02:22	

Data Download Date/Time: 06/26/2019 13:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: MARSHALL

DATE/TIME OF TEST: 06/20/2019 01:15:25

Diagnostics Check		OK	01:09
Air Blank		0.000	01:09
Control Test		0.080	01:09
Air Blank		0.000	01:10
Subject Sample	# 1	0.232	01:11
Breath Volume		1.609	01:11
Air Blank		0.000	01:11
Air Blank		0.000	01:13
Subject Sample	# 2	0.229	01:13
Breath Volume		1.359	01:13
Air Blank		0.000	01:14
Control Test		0.080	01:14
Air Blank		0.000	01:15
Diagnostics Check		OK	01:15

Data Download Date/Time: 06/26/2019 13:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 06/21/2019 01:51:35

Diagnostics Check	OK	01:44
Air Blank	0.000	01:45
Control Test	0.079	01:45
Air Blank	0.000	01:46
Subject Sample # 1	0.087	01:47
Breath Volume	1.109	01:47
Air Blank	0.000	01:47
Air Blank	0.000	01:49
Subject Sample # 2	0.087	01:50
Breath Volume	1.316	01:50
Air Blank	0.000	01:50
Control Test	0.079	01:51
Air Blank	0.000	01:51
Diagnostics Check	OK	01:51

Data Download Date/Time: 06/26/2019 13:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: HERRO

DATE/TIME OF TEST: 06/21/2019 03:10:27

Diagnostics Check	OK	03:07	
Air Blank	0.000	03:08	
Control Test	0.079	03:08	
Air Blank	0.000	03:08	
Subject Sample	# 1	03:09	Subject Test Refused
Breath Volume	0.000	03:09	
Air Blank	0.000	03:09	
Control Test	0.079	03:09	
Air Blank	0.000	03:10	
Diagnostics Check	OK	03:10	

Data Download Date/Time: 06/26/2019 13:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 06/22/2019 00:08:24

Diagnostics Check	OK	00:05	
Air Blank	0.000	00:06	
Control Test	0.080	00:06	
Air Blank	0.000	00:06	
Subject Sample	# 1	00:07	Subject Test Refused
Breath Volume	0.000	00:07	
Air Blank	0.000	00:07	
Control Test	0.080	00:07	
Air Blank	0.000	00:08	
Diagnostics Check	OK	00:08	

Data Download Date/Time: 06/26/2019 13:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: CHESSER

DATE/TIME OF TEST: 06/22/2019 04:10:20

Diagnostics Check	OK	04:04
Air Blank	0.000	04:04
Control Test	0.079	04:04
Air Blank	0.000	04:05
Subject Sample # 1	0.114	04:05
Breath Volume	2.187	04:05
Air Blank	0.000	04:06
Air Blank	0.000	04:08
Subject Sample # 2	0.108	04:08
Breath Volume	1.187	04:08
Air Blank	0.000	04:09
Control Test	0.079	04:09
Air Blank	0.000	04:10
Diagnostics Check	OK	04:10

Data Download Date/Time: 06/26/2019 13:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: BANDA FLORES

DATE/TIME OF TEST: 06/23/2019 22:42:47

Diagnostics Check	OK	22:39	
Air Blank	0.000	22:40	
Control Test	0.081	22:40	
Air Blank	0.000	22:41	
Subject Sample	# 1	22:41	Subject Test Refused
Breath Volume	0.000	22:41	
Air Blank	0.000	22:41	
Control Test	0.079	22:42	
Air Blank	0.000	22:42	
Diagnostics Check	OK	22:42	

Data Download Date/Time: 06/26/2019 13:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: NOBORE,O
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: BILLINGS

DATE/TIME OF TEST: 06/24/2019 00:01:48

Diagnostics Check	OK	23:58	
Air Blank	0.000	23:59	
Control Test	0.079	23:59	
Air Blank	0.000	00:00	
Subject Sample	# 1	00:00	Subject Test Refused
Breath Volume	0.000	00:00	
Air Blank	0.000	00:00	
Control Test	0.079	00:01	
Air Blank	0.000	00:01	
Diagnostics Check	OK	00:01	

Data Download Date/Time: 06/26/2019 13:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: GARVIN

DATE/TIME OF TEST: 06/24/2019 01:51:57

Diagnostics Check	OK	01:45
Air Blank	0.000	01:45
Control Test	0.079	01:46
Air Blank	0.000	01:46
Subject Sample # 1	0.000	01:47
Breath Volume	2.585	01:47
Air Blank	0.000	01:47
Air Blank	0.000	01:49
Subject Sample # 2	0.000	01:50
Breath Volume	3.417	01:50
Air Blank	0.000	01:51
Control Test	0.079	01:51
Air Blank	0.000	01:51
Diagnostics Check	OK	01:51

Data Download Date/Time: 06/26/2019 13:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: PALMER

DATE/TIME OF TEST: 06/26/2019 01:33:07

Diagnostics Check	OK	01:26
Air Blank	0.000	01:27
Control Test	0.079	01:27
Air Blank	0.000	01:28
Subject Sample # 1	0.133	01:28
Breath Volume	1.714	01:28
Air Blank	0.000	01:29
Air Blank	0.000	01:31
Subject Sample # 2	0.131	01:31
Breath Volume	1.871	01:31
Air Blank	0.000	01:32
Control Test	0.079	01:32
Air Blank	0.000	01:33
Diagnostics Check	OK	01:33

Data Download Date/Time: 06/26/2019 13:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: OWENS

DATE/TIME OF TEST: 06/26/2019 22:35:25

Diagnostics Check	OK	22:29
Air Blank	0.000	22:29
Control Test	0.080	22:29
Air Blank	0.000	22:30
Subject Sample # 1	0.055	22:30
Breath Volume	1.324	22:30
Air Blank	0.000	22:31
Air Blank	0.000	22:33
Subject Sample # 2	0.056	22:33
Breath Volume	1.613	22:33
Air Blank	0.000	22:34
Control Test	0.079	22:34
Air Blank	0.000	22:35
Diagnostics Check	OK	22:35

Data Download Date/Time: 07/18/2019 11:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: HOUSTON

DATE/TIME OF TEST: 06/27/2019 04:24:13

Diagnostics Check	OK	04:16
Air Blank	0.000	04:16
Control Test	0.080	04:16
Air Blank	0.000	04:17
Subject Sample # 1	0.000	04:19
Breath Volume	1.218	04:19
Air Blank	0.000	04:20
Air Blank	0.000	04:22
Subject Sample # 2	0.000	04:22
Breath Volume	1.195	04:22
Air Blank	0.000	04:23
Control Test	0.079	04:23
Air Blank	0.000	04:24
Diagnostics Check	OK	04:24

Data Download Date/Time: 07/18/2019 11:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:02

SUBJECT'S LAST NAME: GARDNER

DATE/TIME OF TEST: 06/27/2019 20:26:04

Diagnostics Check	OK	20:23	
Air Blank	0.000	20:23	
Control Test	0.079	20:23	
Air Blank	0.000	20:24	
Subject Sample	# 1	20:24	Subject Test Refused
Breath Volume	0.000	20:24	
Air Blank	0.000	20:25	
Control Test	0.079	20:25	
Air Blank	0.000	20:25	
Diagnostics Check	OK	20:26	

Data Download Date/Time: 07/18/2019 11:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ENGLISH,K
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:51

SUBJECT'S LAST NAME: THIES

DATE/TIME OF TEST: 06/28/2019 05:14:46

Diagnostics Check	OK	05:11	
Air Blank	0.000	05:12	
Control Test	0.080	05:12	
Air Blank	0.000	05:13	
Subject Sample	# 1	05:13	Subject Test Refused
Breath Volume	0.000	05:13	
Air Blank	0.000	05:13	
Control Test	0.080	05:14	
Air Blank	0.000	05:14	
Diagnostics Check	OK	05:14	

Data Download Date/Time: 07/18/2019 11:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: RABIDEAU

DATE/TIME OF TEST: 06/29/2019 01:39:05

Diagnostics Check	OK	01:36	
Air Blank	0.000	01:36	
Control Test	0.079	01:37	
Air Blank	0.000	01:37	
Subject Sample	# 1	01:37	Subject Test Refused
Breath Volume	0.000	01:37	
Air Blank	0.000	01:38	
Control Test	0.080	01:38	
Air Blank	0.000	01:39	
Diagnostics Check	OK	01:39	

Data Download Date/Time: 07/18/2019 11:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: BOYER

DATE/TIME OF TEST: 06/29/2019 04:48:53

Diagnostics Check	OK	04:42
Air Blank	0.000	04:42
Control Test	0.080	04:43
Air Blank	0.000	04:43
Subject Sample # 1	0.203	04:44
Breath Volume	1.109	04:44
Air Blank	0.000	04:45
Air Blank	0.000	04:46
Subject Sample # 2	0.206	04:47
Breath Volume	1.378	04:47
Air Blank	0.000	04:47
Control Test	0.079	04:48
Air Blank	0.000	04:48
Diagnostics Check	OK	04:48

Data Download Date/Time: 07/18/2019 11:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TORRES,W
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: SAVAGE

DATE/TIME OF TEST: 06/30/2019 04:41:47

Diagnostics Check	OK	04:34
Air Blank	0.000	04:35
Control Test	0.080	04:35
Air Blank	0.000	04:36
Subject Sample # 1	0.095	04:37
Breath Volume	1.191	04:37
Air Blank	0.000	04:37
Air Blank	0.000	04:39
Subject Sample # 2	0.101	04:40
Breath Volume	2.105	04:40
Air Blank	0.000	04:40
Control Test	0.079	04:41
Air Blank	0.000	04:41
Diagnostics Check	OK	04:41

Data Download Date/Time: 07/18/2019 11:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001285

INSTRUMENT REGISTERED TO: GILCHRIST COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASH,EARL,C
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MCVEIGH,TREVOR
ARRESTING OFFICER AGENCY: Gilchrist County SO
LAST AGENCY INSPECTION: 05/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1048227
DRY GAS EXPIRY DATE: 11/15/2021
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: WHITTINGTON

DATE/TIME OF TEST: 06/02/2019 01:54:18

Diagnostics Check	OK	01:51	
Air Blank	0.000	01:51	
Control Test	0.081	01:52	
Air Blank	0.000	01:52	
Subject Sample	# 1	01:52	Subject Test Refused
Breath Volume	0.000	01:52	
Air Blank	0.000	01:53	
Control Test	0.081	01:53	
Air Blank	0.000	01:54	
Diagnostics Check	OK	01:54	

Data Download Date/Time: 07/05/2019 13:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001285

INSTRUMENT REGISTERED TO: GILCHRIST COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASH,EARL,C
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MCVEIGH,TREVOR
ARRESTING OFFICER AGENCY: Gilchrist County SO
LAST AGENCY INSPECTION: 05/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1048227
DRY GAS EXPIRY DATE: 11/15/2021
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: WHITTINGTON

DATE/TIME OF TEST: 06/02/2019 02:04:10

Diagnostics Check	OK	02:01	
Air Blank	0.000	02:01	
Control Test	0.080	02:02	
Air Blank	0.000	02:02	
Subject Sample	# 1	02:02	Subject Test Refused
Breath Volume	0.000	02:02	
Air Blank	0.000	02:03	
Control Test	0.080	02:03	
Air Blank	0.000	02:04	
Diagnostics Check	OK	02:04	

Data Download Date/Time: 07/05/2019 13:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001285

INSTRUMENT REGISTERED TO: GILCHRIST COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADKINS,JAKE,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: ADKINS,JAKE
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1048227
DRY GAS EXPIRY DATE: 11/15/2021
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: CRACE

DATE/TIME OF TEST: 06/05/2019 03:43:11

Diagnostics Check	OK	03:40	
Air Blank	0.000	03:40	
Control Test	0.080	03:41	
Air Blank	0.000	03:41	
Subject Sample	# 1	03:41	Subject Test Refused
Breath Volume	0.000	03:41	
Air Blank	0.000	03:42	
Control Test	0.080	03:42	
Air Blank	0.000	03:43	
Diagnostics Check	OK	03:43	

Data Download Date/Time: 07/05/2019 13:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001285

INSTRUMENT REGISTERED TO: GILCHRIST COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADKINS,JAKE,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: ADKINS,JAKE
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1048227
DRY GAS EXPIRY DATE: 11/15/2021
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: CRACE

DATE/TIME OF TEST: 06/05/2019 03:49:20

Diagnostics Check	OK	03:46	
Air Blank	0.000	03:46	
Control Test	0.080	03:47	
Air Blank	0.000	03:47	
Subject Sample	# 1	03:47	Subject Test Refused
Breath Volume	0.000	03:47	
Air Blank	0.000	03:48	
Control Test	0.080	03:48	
Air Blank	0.000	03:49	
Diagnostics Check	OK	03:49	

Data Download Date/Time: 07/05/2019 13:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001285

INSTRUMENT REGISTERED TO: GILCHRIST COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADKINS,JAKE,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: ADKINS,JAKE
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1048227
DRY GAS EXPIRY DATE: 11/15/2021
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: CRACE

DATE/TIME OF TEST: 06/05/2019 03:59:41

Diagnostics Check	OK	03:56	
Air Blank	0.000	03:57	
Control Test	0.080	03:57	
Air Blank	0.000	03:58	
Subject Sample	# 1	03:58	Subject Test Refused
Breath Volume	0.000	03:58	
Air Blank	0.000	03:58	
Control Test	0.080	03:59	
Air Blank	0.000	03:59	
Diagnostics Check	OK	03:59	

Data Download Date/Time: 07/05/2019 13:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001285

INSTRUMENT REGISTERED TO: GILCHRIST COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADKINS, JAKE, M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: VALENCIA, JOVANI
ARRESTING OFFICER AGENCY: Gilchrist County SO
LAST AGENCY INSPECTION: 05/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1048227
DRY GAS EXPIRY DATE: 11/15/2021
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: IORIO

DATE/TIME OF TEST: 06/13/2019 03:03:55

Diagnostics Check	OK	02:56
Air Blank	0.000	02:57
Control Test	0.080	02:57
Air Blank	0.000	02:58
Subject Sample # 1	0.000	02:59
Breath Volume	1.636	02:59
Air Blank	0.000	02:59
Air Blank	0.000	03:01
Subject Sample # 2	0.000	03:02
Breath Volume	2.113	03:02
Air Blank	0.000	03:03
Control Test	0.080	03:03
Air Blank	0.000	03:03
Diagnostics Check	OK	03:03

Data Download Date/Time: 07/05/2019 13:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001285

INSTRUMENT REGISTERED TO: GILCHRIST COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,RICK,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BROWN,RICK
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1048227
DRY GAS EXPIRY DATE: 11/15/2021
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: ROLLING

DATE/TIME OF TEST: 06/16/2019 00:57:09

Diagnostics Check	OK	00:56	
Air Blank		00:56	RFI Detect
Air Blank		00:57	RFI Detect

Data Download Date/Time: 07/05/2019 13:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001285

INSTRUMENT REGISTERED TO: GILCHRIST COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,RICK,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BROWN,RICK
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1048227
DRY GAS EXPIRY DATE: 11/15/2021
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: ROLLING

DATE/TIME OF TEST: 06/16/2019 01:03:43

Diagnostics Check	OK	01:00	
Air Blank	0.000	01:00	
Control Test	0.079	01:01	
Air Blank	0.000	01:01	
Subject Sample # 1	0.261	01:02	
Breath Volume	2.601	01:02	
Air Blank		01:02	RFI Detect
Air Blank	0.000	01:02	
Control Test	0.080	01:03	
Air Blank	0.000	01:03	
Diagnostics Check	OK	01:03	

Data Download Date/Time: 07/05/2019 13:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001285

INSTRUMENT REGISTERED TO: GILCHRIST COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,RICK,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BROWN,RICK
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1048227
DRY GAS EXPIRY DATE: 11/15/2021
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: ROLLING

DATE/TIME OF TEST: 06/16/2019 01:06:57

Diagnostics Check	OK	01:06	
Air Blank		01:06	RFI Detect
Air Blank		01:06	RFI Detect

Data Download Date/Time: 07/05/2019 13:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001285

INSTRUMENT REGISTERED TO: GILCHRIST COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,RICK,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BROWN,RICK
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1048227
DRY GAS EXPIRY DATE: 11/15/2021
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: ROLLING

DATE/TIME OF TEST: 06/16/2019 01:16:23

Diagnostics Check		OK	01:10
Air Blank		0.000	01:10
Control Test		0.079	01:10
Air Blank		0.000	01:11
Subject Sample	# 1	0.258	01:11
Breath Volume		2.910	01:11
Air Blank		0.000	01:12
Air Blank		0.000	01:14
Subject Sample	# 2	0.259	01:14
Breath Volume		2.718	01:14
Air Blank		0.000	01:15
Control Test		0.080	01:15
Air Blank		0.000	01:16
Diagnostics Check		OK	01:16

Data Download Date/Time: 07/05/2019 13:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

INSTRUMENT REGISTERED TO: BAKER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANDTE,A,L
BREATH TEST OPERATOR AGENCY: Baker County SO
ARRESTING OFFICER: FINLEY,M
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: BENDER

DATE/TIME OF TEST: 06/07/2019 03:05:33

Diagnostics Check		OK	02:59
Air Blank		0.000	02:59
Control Test		0.080	02:59
Air Blank		0.000	03:00
Subject Sample	# 1	0.218	03:00
Breath Volume		1.601	03:00
Air Blank		0.000	03:01
Air Blank		0.000	03:03
Subject Sample	# 2	0.209	03:03
Breath Volume		3.484	03:03
Air Blank		0.000	03:04
Control Test		0.079	03:04
Air Blank		0.000	03:05
Diagnostics Check		OK	03:05

Data Download Date/Time: 07/01/2019 20:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HULLENDER,ROBERT,S
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: HULLENDER,ROBERT
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 05/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: TRAYNOR

DATE/TIME OF TEST: 06/11/2019 02:49:07

Diagnostics Check	OK	02:40	
Air Blank	0.000	02:41	
Control Test	0.078	02:41	
Air Blank	0.000	02:41	
Subject Sample	# 1	02:44	No Sample Provided
Breath Volume	0.000	02:44	
Air Blank	0.000	02:45	
Air Blank	0.000	02:47	
Subject Sample	# 2	02:47	Subject Test Refused
Breath Volume	0.000	02:47	
Air Blank	0.000	02:48	
Control Test	0.076	02:48	
Air Blank	0.000	02:49	
Diagnostics Check	OK	02:49	

Data Download Date/Time: 06/21/2019 08:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KONKEL,SCOTT,M
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: TOMPKINS,DANIEL
ARRESTING OFFICER AGENCY: Lawtey PD
LAST AGENCY INSPECTION: 05/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: REAVES

DATE/TIME OF TEST: 06/15/2019 23:42:08

Diagnostics Check	OK	23:35
Air Blank	0.000	23:36
Control Test	0.080	23:36
Air Blank	0.000	23:36
Subject Sample # 1	0.108	23:37
Breath Volume	3.777	23:37
Air Blank	0.000	23:38
Air Blank	0.000	23:39
Subject Sample # 2	0.108	23:40
Breath Volume	3.871	23:40
Air Blank	0.000	23:41
Control Test	0.079	23:41
Air Blank	0.000	23:42
Diagnostics Check	OK	23:42

Data Download Date/Time: 06/21/2019 08:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GORDON,RUSSELL,B
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: HADDOCK,BRIAN
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 05/21/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 04:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 06/20/2019 05:17:47

Diagnostics Check	OK	05:13	
Air Blank	0.000	05:14	
Control Test	0.078	05:14	
Air Blank	0.000	05:15	
Subject Sample # 1	0.127	05:15	
Breath Volume	2.953	05:15	
Air Blank		05:16	RFI Detect
Air Blank	0.000	05:16	
Control Test	0.079	05:17	
Air Blank	0.000	05:17	
Diagnostics Check	OK	05:17	

Data Download Date/Time: 06/21/2019 08:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GORDON,RUSSELL,B
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: HADDOCK,BRIAN
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 05/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 04:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 06/20/2019 05:27:08

Diagnostics Check	OK	05:20
Air Blank	0.000	05:20
Control Test	0.078	05:20
Air Blank	0.000	05:21
Subject Sample # 1	0.124	05:22
Breath Volume	2.363	05:22
Air Blank	0.000	05:23
Air Blank	0.000	05:25
Subject Sample # 2	0.124	05:25
Breath Volume	3.933	05:25
Air Blank	0.000	05:26
Control Test	0.075	05:26
Air Blank	0.000	05:27
Diagnostics Check	OK	05:27

Data Download Date/Time: 06/21/2019 08:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NETTLES, JONATHAN, K
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BEDENBAUGH, JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805701
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:58

SUBJECT'S LAST NAME: KATHLEEN

DATE/TIME OF TEST: 06/01/2019 16:41:08

Diagnostics Check	OK	16:34
Air Blank	0.000	16:35
Control Test	0.079	16:35
Air Blank	0.000	16:36
Subject Sample # 1	0.000	16:36
Breath Volume	3.625	16:36
Air Blank	0.000	16:37
Air Blank	0.000	16:39
Subject Sample # 2	0.000	16:39
Breath Volume	3.242	16:39
Air Blank	0.000	16:40
Control Test	0.077	16:40
Air Blank	0.000	16:41
Diagnostics Check	OK	16:41

Data Download Date/Time: 06/20/2019 19:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLK,AUSTIN,B
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: PETERS,RICH
ARRESTING OFFICER AGENCY: Williston PD
LAST AGENCY INSPECTION: 05/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805701
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: VALLADARES

DATE/TIME OF TEST: 06/04/2019 00:22:49

Diagnostics Check	OK	00:16
Air Blank	0.000	00:16
Control Test	0.079	00:17
Air Blank	0.000	00:17
Subject Sample # 1	0.141	00:18
Breath Volume	1.156	00:18
Air Blank	0.000	00:18
Air Blank	0.000	00:20
Subject Sample # 2	0.135	00:21
Breath Volume	1.144	00:21
Air Blank	0.000	00:21
Control Test	0.078	00:22
Air Blank	0.000	00:22
Diagnostics Check	OK	00:22

Data Download Date/Time: 06/20/2019 19:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADKINS, JAKE, M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: ADKINS, JAKE
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805701
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 06/14/2019 04:55:09

Diagnostics Check	OK	04:48
Air Blank	0.000	04:49
Control Test	0.078	04:49
Air Blank	0.000	04:50
Subject Sample # 1	0.092	04:50
Breath Volume	1.988	04:50
Air Blank	0.000	04:51
Air Blank	0.000	04:53
Subject Sample # 2	0.087	04:53
Breath Volume	1.671	04:53
Air Blank	0.000	04:54
Control Test	0.078	04:54
Air Blank	0.000	04:55
Diagnostics Check	OK	04:55

Data Download Date/Time: 06/20/2019 19:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LESHER, TIFFANY, A
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: DAVIS, KEVIN
ARRESTING OFFICER AGENCY: Levy County SO
LAST AGENCY INSPECTION: 05/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805701
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 12:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:55

SUBJECT'S LAST NAME: SNYDER

DATE/TIME OF TEST: 06/16/2019 13:31:12

Diagnostics Check	OK	13:24
Air Blank	0.000	13:25
Control Test	0.078	13:25
Air Blank	0.000	13:26
Subject Sample # 1	0.188	13:26
Breath Volume	1.582	13:26
Air Blank	0.000	13:27
Air Blank	0.000	13:29
Subject Sample # 2	0.190	13:29
Breath Volume	1.269	13:29
Air Blank	0.000	13:30
Control Test	0.078	13:30
Air Blank	0.000	13:31
Diagnostics Check	OK	13:31

Data Download Date/Time: 06/20/2019 19:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLOOM,PAUL,R
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: GAFFEY,TUCKER
ARRESTING OFFICER AGENCY: Levy County SO
LAST AGENCY INSPECTION: 06/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805701
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 15:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:31

SUBJECT'S LAST NAME: SAMUEL

DATE/TIME OF TEST: 06/27/2019 16:17:00

Diagnostics Check		OK	16:05	
Air Blank		0.000	16:05	
Control Test		0.078	16:05	
Air Blank		0.000	16:06	
Subject Sample	# 1	0.178	16:09	Volume Not Met
Breath Volume		0.550	16:09	
Air Blank		0.000	16:10	
Air Blank		0.000	16:11	
Subject Sample	# 2	0.176	16:12	
Breath Volume		1.285	16:12	
Air Blank		0.000	16:13	
Air Blank		0.000	16:14	
Subject Sample	# 3	0.174	16:15	
Breath Volume		1.656	16:15	
Air Blank		0.000	16:16	
Control Test		0.077	16:16	
Air Blank		0.000	16:16	
Diagnostics Check		OK	16:17	

Data Download Date/Time: 07/31/2019 14:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCONNELL,JAIRUS,S
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MCCONNELL,JAIRUS
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 06/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805701
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: ZERO TOL
ARREST TIME: 14:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:02

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 06/30/2019 15:51:53

Diagnostics Check	OK	15:45
Air Blank	0.000	15:46
Control Test	0.078	15:46
Air Blank	0.000	15:46
Subject Sample # 1	0.050	15:47
Breath Volume	2.042	15:47
Air Blank	0.000	15:48
Air Blank	0.000	15:49
Subject Sample # 2	0.049	15:50
Breath Volume	2.117	15:50
Air Blank	0.000	15:50
Control Test	0.078	15:51
Air Blank	0.000	15:51
Diagnostics Check	OK	15:51

Data Download Date/Time: 07/31/2019 14:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001293

INSTRUMENT REGISTERED TO: WILLISTON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANDIS,BRYAN,S
BREATH TEST OPERATOR AGENCY: Williston PD
ARRESTING OFFICER: LANDIS,BRYAN
ARRESTING OFFICER AGENCY: Williston PD
LAST AGENCY INSPECTION: 05/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 937240
DRY GAS EXPIRY DATE: 04/07/2021
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: SCHUHLEIN

DATE/TIME OF TEST: 06/08/2019 04:16:50

Diagnostics Check	OK	04:10
Air Blank	0.000	04:11
Control Test	0.079	04:11
Air Blank	0.000	04:12
Subject Sample # 1	0.134	04:12
Breath Volume	1.433	04:12
Air Blank	0.000	04:13
Air Blank	0.000	04:14
Subject Sample # 2	0.140	04:15
Breath Volume	2.246	04:15
Air Blank	0.000	04:15
Control Test	0.079	04:16
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

Data Download Date/Time: 06/24/2019 19:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFITH,DALTON,M
BREATH TEST OPERATOR AGENCY: Marianna PD
ARRESTING OFFICER: GRIFFITH,DALTON
ARRESTING OFFICER AGENCY: Marianna PD
LAST AGENCY INSPECTION: 05/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: ROULHAC

DATE/TIME OF TEST: 06/08/2019 01:27:24

Diagnostics Check	OK	01:24	
Air Blank	0.000	01:24	
Control Test	0.082	01:25	
Air Blank	0.000	01:25	
Subject Sample	# 1	01:26	Subject Test Refused
Breath Volume	0.000	01:26	
Air Blank	0.000	01:26	
Control Test	0.082	01:26	
Air Blank	0.000	01:27	
Diagnostics Check	OK	01:27	

Data Download Date/Time: 07/12/2019 15:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARROYO,WILFREDO,
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ARROYO,WILFREDO
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 14:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:15

SUBJECT'S LAST NAME: OSTRANDER

DATE/TIME OF TEST: 06/19/2019 14:47:09

Diagnostics Check	OK	14:39
Air Blank	0.000	14:40
Control Test	0.082	14:40
Air Blank	0.000	14:41
Subject Sample # 1	0.091	14:42
Breath Volume	2.351	14:42
Air Blank	0.000	14:43
Air Blank	0.000	14:45
Subject Sample # 2	0.084	14:45
Breath Volume	2.035	14:45
Air Blank	0.000	14:46
Control Test	0.078	14:46
Air Blank	0.000	14:47
Diagnostics Check	OK	14:47

Data Download Date/Time: 07/12/2019 15:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARROYO,BRANDON,L
BREATH TEST OPERATOR AGENCY: Jackson County SO
ARRESTING OFFICER: ARROYO,TRISTON
ARRESTING OFFICER AGENCY: Marianna PD
LAST AGENCY INSPECTION: 05/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: MALONE

DATE/TIME OF TEST: 06/28/2019 23:20:46

Diagnostics Check	OK	23:15	
Air Blank	0.000	23:16	
Control Test	0.081	23:16	
Air Blank	0.000	23:17	
Subject Sample	# 1	23:19	Subject Test Refused
Breath Volume	0.000	23:19	
Air Blank	0.000	23:19	
Control Test	0.082	23:20	
Air Blank	0.000	23:20	
Diagnostics Check	OK	23:20	

Data Download Date/Time: 07/12/2019 15:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOBEL,EARL,L
BREATH TEST OPERATOR AGENCY: Mexico Beach PD
ARRESTING OFFICER: KNOBEL,EARL
ARRESTING OFFICER AGENCY: Mexico Beach PD
LAST AGENCY INSPECTION: 05/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG722402
DRY GAS EXPIRY DATE: 08/12/2019
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: SCHAMMA

DATE/TIME OF TEST: 06/01/2019 00:12:47

Diagnostics Check	OK	00:06
Air Blank	0.000	00:06
Control Test	0.081	00:07
Air Blank	0.000	00:07
Subject Sample # 1	0.062	00:08
Breath Volume	3.757	00:08
Air Blank	0.000	00:08
Air Blank	0.000	00:10
Subject Sample # 2	0.061	00:11
Breath Volume	3.800	00:11
Air Blank	0.000	00:11
Control Test	0.079	00:12
Air Blank	0.000	00:12
Diagnostics Check	OK	00:12

Data Download Date/Time: 06/28/2019 13:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BASFORD,NICOLE,M
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: BASFORD,NICOLE
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 05/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG722402
DRY GAS EXPIRY DATE: 08/12/2019
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: HUNTLEY

DATE/TIME OF TEST: 06/02/2019 23:19:11

Diagnostics Check	OK	23:18	
Air Blank		23:18	RFI Detect
Air Blank	0.000	23:19	

Data Download Date/Time: 06/28/2019 13:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BASFORD,NICOLE,M
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: BASFORD,NICOLE
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 05/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG722402
DRY GAS EXPIRY DATE: 08/12/2019
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: HUNTLEY

DATE/TIME OF TEST: 06/02/2019 23:28:15

Diagnostics Check	OK	23:21
Air Blank	0.000	23:21
Control Test	0.081	23:21
Air Blank	0.000	23:22
Subject Sample # 1	0.153	23:23
Breath Volume	1.570	23:23
Air Blank	0.000	23:24
Air Blank	0.000	23:26
Subject Sample # 2	0.136	23:26
Breath Volume	2.398	23:26
Air Blank	0.000	23:27
Control Test	0.080	23:27
Air Blank	0.000	23:28
Diagnostics Check	OK	23:28

Data Download Date/Time: 06/28/2019 13:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER,MONTREZ,R
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: POTTER,MONTREZ
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 05/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: ABUNDIS

DATE/TIME OF TEST: 06/02/2019 01:43:17

Diagnostics Check	OK	01:40	
Air Blank	0.000	01:40	
Control Test	0.080	01:41	
Air Blank	0.000	01:41	
Subject Sample	# 1	01:41	Subject Test Refused
Breath Volume	0.000	01:41	
Air Blank	0.000	01:42	
Control Test	0.081	01:42	
Air Blank	0.000	01:43	
Diagnostics Check	OK	01:43	

Data Download Date/Time: 06/28/2019 13:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLYN,WILLIAM,P
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: ALLYN,WILLIAM
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 05/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: HINTON

DATE/TIME OF TEST: 06/04/2019 02:01:29

Diagnostics Check	OK	02:00	
Air Blank		02:01	RFI Detect
Air Blank	0.000	02:01	

Data Download Date/Time: 06/28/2019 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLYN,WILLIAM,P
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: ALLYN,WILLIAM
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 05/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: HINTON

DATE/TIME OF TEST: 06/04/2019 02:08:58

Diagnostics Check	OK	02:02
Air Blank	0.000	02:03
Control Test	0.079	02:03
Air Blank	0.000	02:04
Subject Sample # 1	0.252	02:04
Breath Volume	1.542	02:04
Air Blank	0.000	02:05
Air Blank	0.000	02:07
Subject Sample # 2	0.238	02:07
Breath Volume	1.644	02:07
Air Blank	0.000	02:08
Control Test	0.080	02:08
Air Blank	0.000	02:08
Diagnostics Check	OK	02:08

Data Download Date/Time: 06/28/2019 13:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINCOLN,TREY,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LINCOLN,TREY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 18:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:38

SUBJECT'S LAST NAME: PASANEN

DATE/TIME OF TEST: 06/06/2019 20:06:31

Diagnostics Check	OK	19:58
Air Blank	0.000	19:58
Control Test	0.078	19:59
Air Blank	0.000	19:59
Subject Sample # 1	0.225	20:00
Breath Volume	2.074	20:00
Air Blank	0.000	20:01
Air Blank	0.000	20:03
Subject Sample # 2	0.224	20:04
Breath Volume	2.066	20:04
Air Blank	0.000	20:05
Control Test	0.080	20:05
Air Blank	0.000	20:06
Diagnostics Check	OK	20:06

Data Download Date/Time: 06/28/2019 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER,MONTREZ,R
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: POTTER,MONTREZ
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 05/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: PRINCE

DATE/TIME OF TEST: 06/07/2019 01:20:38

Diagnostics Check	OK	01:14	
Air Blank	0.000	01:14	
Control Test	0.079	01:15	
Air Blank	0.000	01:15	
Subject Sample	# 1 0.305	01:16	
Breath Volume	2.000	01:16	
Air Blank	0.000	01:17	
Air Blank	0.000	01:19	
Subject Sample	# 2	01:19	RFI Detect
Breath Volume	0.000	01:19	
Air Blank	0.000	01:19	
Control Test	0.079	01:20	
Air Blank	0.000	01:20	
Diagnostics Check	OK	01:20	

Data Download Date/Time: 06/28/2019 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER,MONTREZ,R
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: POTTER,MONTREZ
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 05/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: PRINCE

DATE/TIME OF TEST: 06/07/2019 01:28:24

Diagnostics Check	OK	01:22
Air Blank	0.000	01:22
Control Test	0.081	01:22
Air Blank	0.000	01:23
Subject Sample # 1	0.282	01:23
Breath Volume	1.855	01:23
Air Blank	0.000	01:24
Air Blank	0.000	01:26
Subject Sample # 2	0.269	01:26
Breath Volume	1.496	01:26
Air Blank	0.000	01:27
Control Test	0.080	01:27
Air Blank	0.000	01:28
Diagnostics Check	OK	01:28

Data Download Date/Time: 06/28/2019 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,JAMES,J
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: STALNAKER,T
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 05/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 16:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:45

SUBJECT'S LAST NAME: STURGILL

DATE/TIME OF TEST: 06/08/2019 17:20:38

Diagnostics Check	OK	17:13
Air Blank	0.000	17:14
Control Test	0.080	17:14
Air Blank	0.000	17:14
Subject Sample # 1	0.346	17:15
Breath Volume	3.109	17:15
Air Blank	0.000	17:16
Air Blank	0.000	17:18
Subject Sample # 2	0.342	17:19
Breath Volume	3.117	17:19
Air Blank	0.000	17:19
Control Test	0.079	17:20
Air Blank	0.000	17:20
Diagnostics Check	OK	17:20

Data Download Date/Time: 06/28/2019 13:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLYN,WILLIAM,P
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: ALLYN,WILLIAM
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 05/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: TYLER

DATE/TIME OF TEST: 06/14/2019 23:18:34

Diagnostics Check	OK	23:12
Air Blank	0.000	23:12
Control Test	0.080	23:12
Air Blank	0.000	23:13
Subject Sample # 1	0.209	23:14
Breath Volume	2.187	23:14
Air Blank	0.000	23:14
Air Blank	0.000	23:16
Subject Sample # 2	0.215	23:17
Breath Volume	2.250	23:17
Air Blank	0.000	23:17
Control Test	0.079	23:17
Air Blank	0.000	23:18
Diagnostics Check	OK	23:18

Data Download Date/Time: 06/28/2019 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BASFORD,TRAVIS,L
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: BASFORD,TRAVIS
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 05/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: BUI
ARREST TIME: 18:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: LAWLEY

DATE/TIME OF TEST: 06/15/2019 19:08:32

Diagnostics Check	OK	19:02
Air Blank	0.000	19:02
Control Test	0.080	19:03
Air Blank	0.000	19:03
Subject Sample # 1	0.135	19:04
Breath Volume	3.160	19:04
Air Blank	0.000	19:04
Air Blank	0.000	19:06
Subject Sample # 2	0.133	19:07
Breath Volume	3.195	19:07
Air Blank	0.000	19:07
Control Test	0.080	19:07
Air Blank	0.000	19:08
Diagnostics Check	OK	19:08

Data Download Date/Time: 06/28/2019 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDWARDS,KYLE,J
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: BOAN,JULIAN
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 05/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 04:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: MADRIGAL

DATE/TIME OF TEST: 06/16/2019 05:57:34

Diagnostics Check	OK	05:51
Air Blank	0.000	05:51
Control Test	0.081	05:52
Air Blank	0.000	05:52
Subject Sample # 1	0.132	05:53
Breath Volume	2.648	05:53
Air Blank	0.000	05:53
Air Blank	0.000	05:55
Subject Sample # 2	0.127	05:56
Breath Volume	2.097	05:56
Air Blank	0.000	05:56
Control Test	0.080	05:56
Air Blank	0.000	05:57
Diagnostics Check	OK	05:57

Data Download Date/Time: 06/28/2019 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROSIUS,STANLEY,W
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: CICCARELLI,RAY
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 05/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 12:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:05

SUBJECT'S LAST NAME: HAZELIP

DATE/TIME OF TEST: 06/16/2019 13:36:18

Diagnostics Check		OK	13:25
Air Blank		0.000	13:26
Control Test		0.080	13:26
Air Blank		0.000	13:27
Subject Sample	# 1	0.310	13:27
Breath Volume		1.992	13:27
Air Blank		0.000	13:28
Air Blank		0.000	13:30
Subject Sample	# 2	0.288	13:31
Breath Volume		1.964	13:31
Air Blank		0.000	13:32
Air Blank		0.000	13:34
Subject Sample	# 3	0.313	13:34
Breath Volume		1.285	13:34
Air Blank		0.000	13:35
Control Test		0.075	13:35
Air Blank		0.000	13:36
Diagnostics Check		OK	13:36

Data Download Date/Time: 06/28/2019 13:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER,MONTREZ,R
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: POTTER,MONTREZ
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 05/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: KLINGER

DATE/TIME OF TEST: 06/21/2019 23:30:59

Diagnostics Check	OK	23:28	
Air Blank	0.000	23:28	
Control Test	0.080	23:28	
Air Blank	0.000	23:29	
Subject Sample	# 1	23:29	Subject Test Refused
Breath Volume	0.000	23:29	
Air Blank	0.000	23:30	
Control Test	0.078	23:30	
Air Blank	0.000	23:30	
Diagnostics Check	OK	23:30	

Data Download Date/Time: 06/28/2019 13:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER,MONTREZ,R
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: POTTER,MONTREZ
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 06/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 16:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:24

SUBJECT'S LAST NAME: RODRIGUEZ III

DATE/TIME OF TEST: 06/28/2019 17:48:05

Diagnostics Check	OK	17:45	
Air Blank	0.000	17:45	
Control Test	0.075	17:46	
Air Blank	0.000	17:46	
Subject Sample	# 1	17:46	Subject Test Refused
Breath Volume	0.000	17:46	
Air Blank	0.000	17:47	
Control Test	0.078	17:47	
Air Blank	0.000	17:48	
Diagnostics Check	OK	17:48	

Data Download Date/Time: 07/31/2019 13:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSS,CHRISTY,L
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: HUTCHINSON,JOSEPH
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 06/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: CUETO ODUARDO

DATE/TIME OF TEST: 06/29/2019 05:06:37

Diagnostics Check	OK	05:05	
Air Blank		05:06	Ambient Fail
Air Blank	0.000	05:06	

Data Download Date/Time: 07/31/2019 13:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSS,CHRISTY,L
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: HUTCHINSON,JOSEPH
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 06/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:46

SUBJECT'S LAST NAME: CUETO ODUARDO

DATE/TIME OF TEST: 06/29/2019 05:17:34

Diagnostics Check	OK	05:08
Air Blank	0.000	05:09
Control Test	0.079	05:09
Air Blank	0.000	05:10
Subject Sample # 1	0.180	05:13
Breath Volume	2.121	05:13
Air Blank	0.000	05:13
Air Blank	0.000	05:15
Subject Sample # 2	0.178	05:15
Breath Volume	2.300	05:15
Air Blank	0.000	05:16
Control Test	0.077	05:16
Air Blank	0.000	05:17
Diagnostics Check	OK	05:17

Data Download Date/Time: 07/31/2019 13:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRAN,DAVID,V
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: SPRAGUE,CHARLES
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 05/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 171502
DRY GAS EXPIRY DATE: 12/06/2020
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: NEAL

DATE/TIME OF TEST: 06/05/2019 03:20:48

Diagnostics Check	OK	03:17	
Air Blank	0.000	03:17	
Control Test	0.080	03:18	
Air Blank	0.000	03:18	
Subject Sample	# 1	03:19	Subject Test Refused
Breath Volume	0.000	03:19	
Air Blank	0.000	03:19	
Control Test	0.080	03:20	
Air Blank	0.000	03:20	
Diagnostics Check	OK	03:20	

Data Download Date/Time: 07/02/2019 11:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASEY,JOSHUA,P
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: CASEY,JOSHUA
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 05/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 869968
DRY GAS EXPIRY DATE: 12/06/2020
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: LAUDERDALE

DATE/TIME OF TEST: 06/23/2019 02:40:48

Diagnostics Check	OK	02:32
Air Blank	0.000	02:33
Control Test	0.079	02:33
Air Blank	0.000	02:33
Subject Sample # 1	0.098	02:35
Breath Volume	1.175	02:35
Air Blank	0.000	02:36
Air Blank	0.000	02:38
Subject Sample # 2	0.099	02:39
Breath Volume	1.777	02:39
Air Blank	0.000	02:39
Control Test	0.078	02:40
Air Blank	0.000	02:40
Diagnostics Check	OK	02:40

Data Download Date/Time: 07/02/2019 11:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORDHAM,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: RITCHESON,PHILLIP
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 05/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 869968
DRY GAS EXPIRY DATE: 12/06/2020
VIOLATION CODE: DUI
ARREST TIME: 09:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:27

SUBJECT'S LAST NAME: KENNEDY

DATE/TIME OF TEST: 06/25/2019 12:50:34

Diagnostics Check	OK	12:47	
Air Blank	0.000	12:48	
Control Test	0.079	12:48	
Air Blank	0.000	12:48	
Subject Sample	# 1	12:49	Subject Test Refused
Breath Volume	0.000	12:49	
Air Blank	0.000	12:49	
Control Test	0.079	12:49	
Air Blank	0.000	12:50	
Diagnostics Check	OK	12:50	

Data Download Date/Time: 07/02/2019 11:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNGHANSE,KYLE,O
BREATH TEST OPERATOR AGENCY: DeFuniak Springs PD
ARRESTING OFFICER: YOUNGHANSE,KYLE
ARRESTING OFFICER AGENCY: DeFuniak Springs PD
LAST AGENCY INSPECTION: 05/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 05:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: JIMENEZ

DATE/TIME OF TEST: 06/02/2019 06:15:54

Diagnostics Check	OK	06:08
Air Blank	0.000	06:09
Control Test	0.080	06:09
Air Blank	0.000	06:10
Subject Sample # 1	0.176	06:11
Breath Volume	1.468	06:11
Air Blank	0.000	06:11
Air Blank	0.000	06:13
Subject Sample # 2	0.169	06:14
Breath Volume	2.566	06:14
Air Blank	0.000	06:14
Control Test	0.079	06:15
Air Blank	0.000	06:15
Diagnostics Check	OK	06:15

Data Download Date/Time: 06/25/2019 03:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPANN,BRENNEN,W
BREATH TEST OPERATOR AGENCY: DeFuniak Springs PD
ARRESTING OFFICER: SPANN,BRENNEN
ARRESTING OFFICER AGENCY: DeFuniak Springs PD
LAST AGENCY INSPECTION: 05/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 19:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: BRADSHAW

DATE/TIME OF TEST: 06/02/2019 22:38:10

Diagnostics Check	OK	22:26	
Air Blank	0.000	22:27	
Control Test	0.081	22:27	
Air Blank	0.000	22:28	
Subject Sample	# 1	22:31	No Sample Provided
Breath Volume	0.000	22:31	
Air Blank	0.000	22:31	
Air Blank	0.000	22:33	
Subject Sample	# 2	22:36	Volume Not Met
Breath Volume	0.355	22:36	
Air Blank	0.000	22:37	
Control Test	0.082	22:37	
Air Blank	0.000	22:38	
Diagnostics Check	OK	22:38	

Data Download Date/Time: 06/25/2019 03:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENNETT,DAVID,J
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: LOCKLEAR,SARGE
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 05/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: OTHER
ARREST TIME: 10:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:50

SUBJECT'S LAST NAME: TYUS

DATE/TIME OF TEST: 06/08/2019 11:39:23

Diagnostics Check	OK	11:33	
Air Blank	0.000	11:34	
Control Test	0.080	11:34	
Air Blank	0.000	11:34	
Subject Sample # 1	0.027	11:35	
Breath Volume	2.628	11:35	
Air Blank	0.000	11:35	
Air Blank	0.000	11:37	
		11:37	Improper Sample
Air Blank	0.000	11:38	
Control Test	0.083	11:38	
Air Blank	0.000	11:39	
Diagnostics Check	OK	11:39	

Data Download Date/Time: 06/25/2019 03:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITAKER,JOHNNY,W
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WHITAKER,JOHNNY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: FINK

DATE/TIME OF TEST: 06/12/2019 03:19:29

Diagnostics Check	OK	03:15	
Air Blank	0.000	03:15	
Control Test	0.080	03:16	
Air Blank	0.000	03:16	
Subject Sample	# 1	03:19	Sequence Aborted
Breath Volume	0.000	03:19	
Air Blank	0.000	03:19	

Data Download Date/Time: 06/25/2019 03:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITAKER,JOHNNY,W
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WHITAKER,JOHNNY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: FINK

DATE/TIME OF TEST: 06/12/2019 03:33:44

Diagnostics Check	OK	03:22	
Air Blank	0.000	03:22	
Control Test	0.080	03:23	
Air Blank	0.000	03:23	
Subject Sample	# 1	03:26	No Sample Provided
Breath Volume	0.000	03:26	
Air Blank	0.000	03:27	
Air Blank	0.000	03:29	
Subject Sample	# 2	03:32	No Sample Provided
Breath Volume	0.000	03:32	
Air Blank	0.000	03:32	
Control Test	0.082	03:33	
Air Blank	0.000	03:33	
Diagnostics Check	OK	03:33	

Data Download Date/Time: 06/25/2019 03:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,DAVID,O
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: MILLER,WESLEY
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 05/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 07:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:04

SUBJECT'S LAST NAME: STRICKLAND

DATE/TIME OF TEST: 06/12/2019 08:43:21

Diagnostics Check	OK	08:31
Air Blank	0.000	08:32
Control Test	0.080	08:32
Air Blank	0.000	08:32
Subject Sample # 1	0.142	08:36
Breath Volume	2.675	08:36
Air Blank	0.000	08:36
Air Blank	0.000	08:38
Subject Sample # 2	0.138	08:41
Breath Volume	1.789	08:41
Air Blank	0.000	08:42
Control Test	0.079	08:42
Air Blank	0.000	08:43
Diagnostics Check	OK	08:43

Data Download Date/Time: 06/25/2019 03:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CURTIS,WILLIAM,E
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: TICE,HUNTER
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 05/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 13:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:22

SUBJECT'S LAST NAME: DEMPSEY

DATE/TIME OF TEST: 06/17/2019 15:55:26

Diagnostics Check	OK	15:49
Air Blank	0.000	15:49
Control Test	0.079	15:50
Air Blank	0.000	15:50
Subject Sample # 1	0.000	15:51
Breath Volume	2.257	15:51
Air Blank	0.000	15:51
Air Blank	0.000	15:53
Subject Sample # 2	0.000	15:54
Breath Volume	1.554	15:54
Air Blank	0.000	15:54
Control Test	0.080	15:54
Air Blank	0.000	15:55
Diagnostics Check	OK	15:55

Data Download Date/Time: 06/25/2019 03:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUNCEY,KATHLEEN,R
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: POUNCEY,KATHLEEN
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 05/30/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 09:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:45

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 06/19/2019 12:18:55

Diagnostics Check	OK	12:12
Air Blank	0.000	12:12
Control Test	0.080	12:13
Air Blank	0.000	12:13
Subject Sample # 1	0.067	12:14
Breath Volume	1.707	12:14
Air Blank	0.000	12:14
Air Blank	0.000	12:16
Subject Sample # 2	0.062	12:17
Breath Volume	2.035	12:17
Air Blank	0.000	12:18
Control Test	0.080	12:18
Air Blank	0.000	12:18
Diagnostics Check	OK	12:18

Data Download Date/Time: 06/25/2019 03:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHULTZ,JEREMY,J
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: SCHULTZ,JEREMY
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 05/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: MAYER

DATE/TIME OF TEST: 06/22/2019 01:36:09

Diagnostics Check	OK	01:24	
Air Blank	0.000	01:25	
Control Test	0.080	01:25	
Air Blank	0.000	01:26	
Subject Sample	# 1	01:29	No Sample Provided
Breath Volume	0.000	01:29	
Air Blank	0.000	01:29	
Air Blank	0.000	01:31	
Subject Sample	# 2	01:34	No Sample Provided
Breath Volume	0.000	01:34	
Air Blank	0.000	01:35	
Control Test	0.081	01:35	
Air Blank	0.000	01:36	
Diagnostics Check	OK	01:36	

Data Download Date/Time: 06/25/2019 03:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOBLES,DARIUS,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: NOBLES,DARIUS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 17:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: BURNS

DATE/TIME OF TEST: 06/22/2019 18:57:58

Diagnostics Check	OK	18:49
Air Blank	0.000	18:49
Control Test	0.080	18:50
Air Blank	0.000	18:50
Subject Sample # 1	0.266	18:52
Breath Volume	3.417	18:52
Air Blank	0.000	18:53
Air Blank	0.000	18:55
Subject Sample # 2	0.276	18:56
Breath Volume	3.699	18:56
Air Blank	0.000	18:57
Control Test	0.078	18:57
Air Blank	0.000	18:57
Diagnostics Check	OK	18:57

Data Download Date/Time: 06/25/2019 03:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCLEOD,JOEL,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MCLEOD,JOEL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: SHEPPARD

DATE/TIME OF TEST: 06/23/2019 02:19:14

Diagnostics Check		OK	02:12
Air Blank		0.000	02:12
Control Test		0.081	02:12
Air Blank		0.000	02:13
Subject Sample	# 1	0.208	02:13
Breath Volume		1.230	02:13
Air Blank		0.000	02:14
Air Blank		0.000	02:16
Subject Sample	# 2	0.209	02:17
Breath Volume		2.507	02:17
Air Blank		0.000	02:18
Control Test		0.081	02:18
Air Blank		0.000	02:19
Diagnostics Check		OK	02:19

Data Download Date/Time: 06/25/2019 03:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON, BRANDON, W
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK, STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: G620013897071

DATE/TIME OF TEST: 06/01/2019 03:27:19

Diagnostics Check	OK	03:21
Air Blank	0.000	03:21
Control Test	0.081	03:21
Air Blank	0.000	03:22
Subject Sample # 1	0.169	03:22
Breath Volume	3.203	03:22
Air Blank	0.000	03:23
Air Blank	0.000	03:25
Subject Sample # 2	0.166	03:25
Breath Volume	2.890	03:25
Air Blank	0.000	03:26
Control Test	0.080	03:26
Air Blank	0.000	03:27
Diagnostics Check	OK	03:27

Data Download Date/Time: 06/04/2019 10:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,STEVEN,D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK,STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 06/01/2019 23:20:32

Diagnostics Check	OK	23:17	
Air Blank	0.000	23:18	
Control Test	0.079	23:18	
Air Blank	0.000	23:18	
Subject Sample	# 1	23:19	Subject Test Refused
Breath Volume	0.000	23:19	
Air Blank	0.000	23:19	
Control Test	0.079	23:19	
Air Blank	0.000	23:20	
Diagnostics Check	OK	23:20	

Data Download Date/Time: 06/04/2019 10:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON, BRANDON, W
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: JONES, CARL
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: SOLIZ

DATE/TIME OF TEST: 06/02/2019 01:19:17

Diagnostics Check	OK	01:10	
Air Blank	0.000	01:11	
Control Test	0.081	01:11	
Air Blank	0.000	01:11	
Subject Sample	# 1	01:15	No Sample Provided
Breath Volume	0.000	01:15	
Air Blank	0.000	01:15	
Air Blank	0.000	01:17	
Subject Sample	# 2	01:17	Subject Test Refused
Breath Volume	0.000	01:17	
Air Blank	0.000	01:18	
Control Test	0.080	01:18	
Air Blank	0.000	01:19	
Diagnostics Check	OK	01:19	

Data Download Date/Time: 06/04/2019 10:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON, BRANDON, W
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ARISTY, IVAN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: PEARCE

DATE/TIME OF TEST: 06/02/2019 01:46:11

Diagnostics Check	OK	01:43	
Air Blank	0.000	01:43	
Control Test	0.080	01:43	
Air Blank	0.000	01:44	
Subject Sample	# 1	01:44	Subject Test Refused
Breath Volume	0.000	01:44	
Air Blank	0.000	01:45	
Control Test	0.081	01:45	
Air Blank	0.000	01:46	
Diagnostics Check	OK	01:46	

Data Download Date/Time: 06/04/2019 10:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT, MARK, A
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK, S
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 06/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: PONCE

DATE/TIME OF TEST: 06/09/2019 02:06:36

Diagnostics Check	OK	02:03	
Air Blank	0.000	02:04	
Control Test	0.080	02:04	
Air Blank	0.000	02:04	
Subject Sample	# 1	02:05	Subject Test Refused
Breath Volume	0.000	02:05	
Air Blank	0.000	02:05	
Control Test	0.080	02:06	
Air Blank	0.000	02:06	
Diagnostics Check	OK	02:06	

Data Download Date/Time: 07/02/2019 13:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,MARK,A
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: VALETA,B
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 06/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: FARR

DATE/TIME OF TEST: 06/09/2019 03:05:34

Diagnostics Check	OK	02:59
Air Blank	0.000	02:59
Control Test	0.081	03:00
Air Blank	0.000	03:00
Subject Sample # 1	0.195	03:01
Breath Volume	1.605	03:01
Air Blank	0.000	03:01
Air Blank	0.000	03:03
Subject Sample # 2	0.197	03:04
Breath Volume	1.476	03:04
Air Blank	0.000	03:04
Control Test	0.079	03:04
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

Data Download Date/Time: 07/02/2019 13:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,MARK,A
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: VULETA,B
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 06/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 06/13/2019 03:29:38

Diagnostics Check	OK	03:23
Air Blank	0.000	03:23
Control Test	0.080	03:24
Air Blank	0.000	03:24
Subject Sample # 1	0.195	03:25
Breath Volume	1.660	03:25
Air Blank	0.000	03:25
Air Blank	0.000	03:27
Subject Sample # 2	0.199	03:28
Breath Volume	2.585	03:28
Air Blank	0.000	03:28
Control Test	0.080	03:29
Air Blank	0.000	03:29
Diagnostics Check	OK	03:29

Data Download Date/Time: 07/02/2019 13:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALCAGNO, MERCEDES, T
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: VELETTA, BENJAMIN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 06/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 19:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: YOTHER

DATE/TIME OF TEST: 06/16/2019 20:15:33

Diagnostics Check	OK	20:09
Air Blank	0.000	20:09
Control Test	0.079	20:10
Air Blank	0.000	20:10
Subject Sample # 1	0.262	20:10
Breath Volume	3.917	20:10
Air Blank	0.000	20:11
Air Blank	0.000	20:13
Subject Sample # 2	0.262	20:13
Breath Volume	2.933	20:13
Air Blank	0.000	20:14
Control Test	0.076	20:14
Air Blank	0.000	20:15
Diagnostics Check	OK	20:15

Data Download Date/Time: 07/02/2019 13:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALCAGNO, MERCEDES, T
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: VULETA, BENJAMIN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 06/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:03

SUBJECT'S LAST NAME: GARCIAGONZALEZ

DATE/TIME OF TEST: 06/16/2019 21:32:58

Diagnostics Check	OK	21:30	
Air Blank	0.000	21:30	
Control Test	0.080	21:30	
Air Blank	0.000	21:31	
Subject Sample	# 1	21:31	Subject Test Refused
Breath Volume	0.000	21:31	
Air Blank	0.000	21:32	
Control Test	0.078	21:32	
Air Blank	0.000	21:32	
Diagnostics Check	OK	21:32	

Data Download Date/Time: 07/02/2019 13:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,MARK,A
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: VULETA,B
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 06/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: DIXON

DATE/TIME OF TEST: 06/18/2019 01:50:59

Diagnostics Check	OK	01:44
Air Blank	0.000	01:45
Control Test	0.080	01:45
Air Blank	0.000	01:45
Subject Sample # 1	0.269	01:46
Breath Volume	1.675	01:46
Air Blank	0.000	01:46
Air Blank	0.000	01:48
Subject Sample # 2	0.253	01:49
Breath Volume	3.023	01:49
Air Blank	0.000	01:50
Control Test	0.079	01:50
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

Data Download Date/Time: 07/02/2019 13:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAUSLER,TREVOR,C
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: HAUSLER,TREVOR
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 06/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: BUI
ARREST TIME: 19:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:51

SUBJECT'S LAST NAME: SIERACKI

DATE/TIME OF TEST: 06/22/2019 21:14:37

Diagnostics Check	OK	21:11	
Air Blank	0.000	21:12	
Control Test	0.080	21:12	
Air Blank	0.000	21:12	
Subject Sample	# 1	21:13	Subject Test Refused
Breath Volume	0.000	21:13	
Air Blank	0.000	21:13	
Control Test	0.080	21:14	
Air Blank	0.000	21:14	
Diagnostics Check	OK	21:14	

Data Download Date/Time: 07/02/2019 13:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT, MARK, A
BREATH TEST OPERATOR AGENCY: Okeechobee PD
ARRESTING OFFICER: FRANCIS, J
ARRESTING OFFICER AGENCY: Okeechobee PD
LAST AGENCY INSPECTION: 06/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: LOWE

DATE/TIME OF TEST: 06/23/2019 03:25:46

Diagnostics Check	OK	03:22	
Air Blank	0.000	03:23	
Control Test	0.080	03:23	
Air Blank	0.000	03:24	
Subject Sample	# 1	03:24	Subject Test Refused
Breath Volume	0.000	03:24	
Air Blank	0.000	03:24	
Control Test	0.080	03:25	
Air Blank	0.000	03:25	
Diagnostics Check	OK	03:25	

Data Download Date/Time: 07/02/2019 13:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,STEVEN,D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK,STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 06/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: SHEPHERD

DATE/TIME OF TEST: 06/29/2019 01:49:12

Diagnostics Check	OK	01:42
Air Blank	0.000	01:43
Control Test	0.080	01:43
Air Blank	0.000	01:43
Subject Sample # 1	0.213	01:44
Breath Volume	3.500	01:44
Air Blank	0.000	01:45
Air Blank	0.000	01:46
Subject Sample # 2	0.196	01:47
Breath Volume	2.511	01:47
Air Blank	0.000	01:48
Control Test	0.080	01:48
Air Blank	0.000	01:49
Diagnostics Check	OK	01:49

Data Download Date/Time: 07/02/2019 13:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,STEVEN,D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK,STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 06/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 06/30/2019 01:50:00

Diagnostics Check	OK	01:43
Air Blank	0.000	01:43
Control Test	0.080	01:44
Air Blank	0.000	01:44
Subject Sample # 1	0.133	01:45
Breath Volume	2.601	01:45
Air Blank	0.000	01:46
Air Blank	0.000	01:47
Subject Sample # 2	0.134	01:48
Breath Volume	1.875	01:48
Air Blank	0.000	01:49
Control Test	0.079	01:49
Air Blank	0.000	01:49
Diagnostics Check	OK	01:50

Data Download Date/Time: 07/02/2019 13:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON, BRANDON, W
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: HAYLES, SHANE
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 06/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 19:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: OLSON

DATE/TIME OF TEST: 06/30/2019 20:33:11

Diagnostics Check	OK	20:30	
Air Blank	0.000	20:30	
Control Test	0.080	20:30	
Air Blank	0.000	20:31	
Subject Sample	# 1	20:31	Subject Test Refused
Breath Volume	0.000	20:31	
Air Blank	0.000	20:32	
Control Test	0.080	20:32	
Air Blank	0.000	20:33	
Diagnostics Check	OK	20:33	

Data Download Date/Time: 07/02/2019 13:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TREVISOL,JOSEPH,D
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: TREVISOL,JOSEPH
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 05/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 15:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:50

SUBJECT'S LAST NAME: DURHAM

DATE/TIME OF TEST: 06/01/2019 16:17:45

Diagnostics Check	OK	16:10
Air Blank	0.000	16:11
Control Test	0.081	16:11
Air Blank	0.000	16:12
Subject Sample # 1	0.187	16:12
Breath Volume	2.304	16:12
Air Blank	0.000	16:13
Air Blank	0.000	16:15
Subject Sample # 2	0.190	16:16
Breath Volume	2.531	16:16
Air Blank	0.000	16:16
Control Test	0.080	16:17
Air Blank	0.000	16:17
Diagnostics Check	OK	16:17

Data Download Date/Time: 06/29/2019 18:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OJEDA,NELSON,J
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: OJEDA,NELSON
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 05/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: MCMAHON

DATE/TIME OF TEST: 06/16/2019 03:58:38

Diagnostics Check		OK	03:51
Air Blank		0.000	03:51
Control Test		0.081	03:52
Air Blank		0.000	03:52
Subject Sample	# 1	0.204	03:53
Breath Volume		1.960	03:53
Air Blank		0.000	03:54
Air Blank		0.000	03:56
Subject Sample	# 2	0.195	03:57
Breath Volume		1.527	03:57
Air Blank		0.000	03:57
Control Test		0.079	03:58
Air Blank		0.000	03:58
Diagnostics Check		OK	03:58

Data Download Date/Time: 06/29/2019 18:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEARSE,ANGIE,D
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: DELEON,TEA
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 05/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 06/01/2019 02:10:01

Diagnostics Check	OK	02:02
Air Blank	0.000	02:02
Control Test	0.079	02:03
Air Blank	0.000	02:03
Subject Sample # 1	0.292	02:05
Breath Volume	1.363	02:05
Air Blank	0.000	02:06
Air Blank	0.000	02:07
Subject Sample # 2	0.296	02:08
Breath Volume	2.878	02:08
Air Blank	0.000	02:09
Control Test	0.079	02:09
Air Blank	0.000	02:09
Diagnostics Check	OK	02:10

Data Download Date/Time: 06/03/2019 02:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANDY,KEVIN,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: VAVRA, TYLER
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 05/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: DINUNZIO

DATE/TIME OF TEST: 06/02/2019 03:44:42

Diagnostics Check	OK	03:41	
Air Blank	0.000	03:42	
Control Test	0.079	03:42	
Air Blank	0.000	03:43	
Subject Sample	# 1	03:43	Subject Test Refused
Breath Volume	0.000	03:43	
Air Blank	0.000	03:43	
Control Test	0.080	03:44	
Air Blank	0.000	03:44	
Diagnostics Check	OK	03:44	

Data Download Date/Time: 06/03/2019 02:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FONSECA,GIOVANNY,
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: FONSECA,GIOVANNY
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 05/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: BAUGHMAN

DATE/TIME OF TEST: 06/02/2019 04:47:40

Diagnostics Check	OK	04:40
Air Blank	0.000	04:41
Control Test	0.080	04:41
Air Blank	0.000	04:42
Subject Sample # 1	0.168	04:42
Breath Volume	1.609	04:42
Air Blank	0.000	04:43
Air Blank	0.000	04:45
Subject Sample # 2	0.174	04:46
Breath Volume	1.781	04:46
Air Blank	0.000	04:46
Control Test	0.079	04:47
Air Blank	0.000	04:47
Diagnostics Check	OK	04:47

Data Download Date/Time: 06/03/2019 02:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MAEGAN, R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE, B
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 05/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 16:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: SANTIAGO TORRES

DATE/TIME OF TEST: 06/02/2019 18:50:01

Diagnostics Check	OK	18:47	
Air Blank	0.000	18:47	
Control Test	0.080	18:47	
Air Blank	0.000	18:48	
Subject Sample	# 1	18:48	Subject Test Refused
Breath Volume	0.000	18:48	
Air Blank	0.000	18:49	
Control Test	0.079	18:49	
Air Blank	0.000	18:49	
Diagnostics Check	OK	18:50	

Data Download Date/Time: 06/03/2019 02:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAYMOND,FRANK,W
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: FLORES,DAVID
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 05/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: JIMENEZ

DATE/TIME OF TEST: 06/04/2019 03:58:45

Diagnostics Check	OK	03:55	
Air Blank	0.000	03:56	
Control Test	0.080	03:56	
Air Blank	0.000	03:57	
Subject Sample	# 1	03:57	Subject Test Refused
Breath Volume	0.000	03:57	
Air Blank	0.000	03:57	
Control Test	0.080	03:58	
Air Blank	0.000	03:58	
Diagnostics Check	OK	03:58	

Data Download Date/Time: 06/10/2019 19:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIGMORELLI,A,W
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: CASTANO,C
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 05/31/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 21:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: HANKS

DATE/TIME OF TEST: 06/04/2019 22:25:24

Diagnostics Check	OK	22:18
Air Blank	0.000	22:19
Control Test	0.079	22:19
Air Blank	0.000	22:20
Subject Sample # 1	0.055	22:20
Breath Volume	3.558	22:20
Air Blank	0.000	22:21
Air Blank	0.000	22:23
Subject Sample # 2	0.052	22:23
Breath Volume	2.613	22:23
Air Blank	0.000	22:24
Control Test	0.079	22:24
Air Blank	0.000	22:25
Diagnostics Check	OK	22:25

Data Download Date/Time: 06/10/2019 19:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLIDAY,LATONYA,A
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: VAVIA,T.
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 05/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 06/07/2019 04:08:35

Diagnostics Check	OK	04:05	
Air Blank	0.000	04:06	
Control Test	0.080	04:06	
Air Blank	0.000	04:07	
Subject Sample	# 1	04:07	Subject Test Refused
Breath Volume	0.000	04:07	
Air Blank	0.000	04:07	
Control Test	0.080	04:08	
Air Blank	0.000	04:08	
Diagnostics Check	OK	04:08	

Data Download Date/Time: 06/10/2019 19:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FONSECA,GIOVANNY,
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: FONSECA,GIOVANNY
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 06/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: LOPEZ MORALES

DATE/TIME OF TEST: 06/15/2019 02:22:21

Diagnostics Check		OK	02:15
Air Blank		0.000	02:16
Control Test		0.079	02:16
Air Blank		0.000	02:16
Subject Sample	# 1	0.223	02:17
Breath Volume		1.476	02:17
Air Blank		0.000	02:18
Air Blank		0.000	02:20
Subject Sample	# 2	0.222	02:20
Breath Volume		2.367	02:20
Air Blank		0.000	02:21
Control Test		0.078	02:21
Air Blank		0.000	02:22
Diagnostics Check		OK	02:22

Data Download Date/Time: 07/25/2019 05:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAYMOND,FRANK,W
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: CASTANO,C
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 06/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: EDMOND

DATE/TIME OF TEST: 06/18/2019 00:33:07

Diagnostics Check	OK	00:26
Air Blank	0.000	00:26
Control Test	0.080	00:27
Air Blank	0.000	00:27
Subject Sample # 1	0.010	00:28
Breath Volume	1.539	00:28
Air Blank	0.000	00:28
Air Blank	0.000	00:30
Subject Sample # 2	0.010	00:31
Breath Volume	1.910	00:31
Air Blank	0.000	00:32
Control Test	0.080	00:32
Air Blank	0.000	00:33
Diagnostics Check	OK	00:33

Data Download Date/Time: 07/25/2019 05:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIGNORELLIA,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: ROLLANDO,T
ARRESTING OFFICER AGENCY: Fellsmere PD
LAST AGENCY INSPECTION: 06/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:23

SUBJECT'S LAST NAME: TWORKOWSKI

DATE/TIME OF TEST: 06/20/2019 22:53:05

Diagnostics Check	OK	22:50	
Air Blank	0.000	22:50	
Control Test	0.080	22:50	
Air Blank	0.000	22:51	
Subject Sample	# 1	22:51	Subject Test Refused
Breath Volume	0.000	22:51	
Air Blank	0.000	22:52	
Control Test	0.080	22:52	
Air Blank	0.000	22:53	
Diagnostics Check	OK	22:53	

Data Download Date/Time: 07/25/2019 05:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANDY,KEVIN,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: SANDERS,BRIDGETTE
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 06/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: HAASE-HUGHES

DATE/TIME OF TEST: 06/21/2019 02:48:56

Diagnostics Check	OK	02:46	
Air Blank	0.000	02:46	
Control Test	0.079	02:46	
Air Blank	0.000	02:47	
Subject Sample	# 1	02:47	Subject Test Refused
Breath Volume	0.000	02:47	
Air Blank	0.000	02:48	
Control Test	0.079	02:48	
Air Blank	0.000	02:48	
Diagnostics Check	OK	02:48	

Data Download Date/Time: 07/25/2019 05:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HISS,JEFFREY,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: WHITTINGTON,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 06/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 18:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: BRODBECK

DATE/TIME OF TEST: 06/22/2019 19:58:20

Diagnostics Check	OK	19:55	
Air Blank	0.000	19:55	
Control Test	0.079	19:56	
Air Blank	0.000	19:56	
Subject Sample	# 1	19:56	Subject Test Refused
Breath Volume	0.000	19:56	
Air Blank	0.000	19:57	
Control Test	0.080	19:57	
Air Blank	0.000	19:58	
Diagnostics Check	OK	19:58	

Data Download Date/Time: 07/25/2019 05:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEARSE,ANGIE,D
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,DENNIS
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 06/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: MEADOWS

DATE/TIME OF TEST: 06/23/2019 01:42:30

Diagnostics Check	OK	01:41	
Air Blank	0.000	01:42	
		01:42	RFI Detect
Air Blank		01:42	RFI Detect

Data Download Date/Time: 07/25/2019 05:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,QUINTON,T
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: FRICKE,JOANNA
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 06/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 06/23/2019 23:42:12

Diagnostics Check	OK	23:36
Air Blank	0.000	23:36
Control Test	0.080	23:36
Air Blank	0.000	23:37
Subject Sample # 1	0.180	23:37
Breath Volume	1.789	23:37
Air Blank	0.000	23:38
Air Blank	0.000	23:40
Subject Sample # 2	0.187	23:40
Breath Volume	1.835	23:40
Air Blank	0.000	23:41
Control Test	0.079	23:41
Air Blank	0.000	23:42
Diagnostics Check	OK	23:42

Data Download Date/Time: 07/25/2019 05:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHIMENTI,RICHARD,J
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: SCROGGIN,JOHN
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 06/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 16:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: FAULKNER

DATE/TIME OF TEST: 06/25/2019 17:26:28

Diagnostics Check	OK	17:21	
Air Blank	0.000	17:21	
Control Test	0.079	17:22	
Air Blank	0.000	17:22	
Subject Sample	# 1	17:25	Subject Test Refused
Breath Volume	0.000	17:25	
Air Blank	0.000	17:25	
Control Test	0.079	17:25	
Air Blank	0.000	17:26	
Diagnostics Check	OK	17:26	

Data Download Date/Time: 07/25/2019 05:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KANTORSKI,MARK,A
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: SINGH,TEGPREET
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 06/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: RUSSELL

DATE/TIME OF TEST: 06/26/2019 02:45:51

Diagnostics Check	OK	02:38
Air Blank	0.000	02:39
Control Test	0.079	02:39
Air Blank	0.000	02:39
Subject Sample # 1	0.137	02:40
Breath Volume	3.039	02:40
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.135	02:44
Breath Volume	2.460	02:44
Air Blank	0.000	02:44
Control Test	0.080	02:45
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 07/25/2019 05:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLIDAY,LATONYA,A
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: HATTON,N
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 06/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 06/26/2019 05:25:39

Diagnostics Check	OK	05:23	
Air Blank	0.000	05:24	
Control Test	0.080	05:24	
Air Blank	0.000	05:25	
Subject Sample	# 1	05:25	Sequence Aborted
Breath Volume	0.000	05:25	
Air Blank	0.000	05:25	

Data Download Date/Time: 07/25/2019 05:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,TERRENCE,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: HATTON,N
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 06/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 04:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:42

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 06/26/2019 09:08:46

Diagnostics Check	OK	09:02
Air Blank	0.000	09:02
Control Test	0.079	09:03
Air Blank	0.000	09:03
Subject Sample # 1	0.063	09:04
Breath Volume	3.828	09:04
Air Blank	0.000	09:04
Air Blank	0.000	09:06
Subject Sample # 2	0.061	09:07
Breath Volume	3.125	09:07
Air Blank	0.000	09:07
Control Test	0.080	09:08
Air Blank	0.000	09:08
Diagnostics Check	OK	09:08

Data Download Date/Time: 07/25/2019 05:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWINDELL,CHRISTOPHER,E
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: RUIZ,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 06/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: MARMOLEJO-MENDOZA

DATE/TIME OF TEST: 06/27/2019 01:13:28

Diagnostics Check	OK	01:10	
Air Blank	0.000	01:11	
Control Test	0.079	01:11	
Air Blank	0.000	01:11	
Subject Sample	# 1	01:12	Subject Test Refused
Breath Volume	0.000	01:12	
Air Blank	0.000	01:12	
Control Test	0.079	01:12	
Air Blank	0.000	01:13	
Diagnostics Check	OK	01:13	

Data Download Date/Time: 07/25/2019 05:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHILE,C,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,B
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 06/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 18:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: HAGAN

DATE/TIME OF TEST: 06/27/2019 19:57:22

Diagnostics Check	OK	19:54	
Air Blank	0.000	19:54	
Control Test	0.080	19:55	
Air Blank	0.000	19:55	
Subject Sample	# 1	19:56	Subject Test Refused
Breath Volume	0.000	19:56	
Air Blank	0.000	19:56	
Control Test	0.079	19:56	
Air Blank		19:56	RFI Detect
Air Blank	0.000	19:57	

Data Download Date/Time: 07/25/2019 05:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIGNORELLIA,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: SINGH,T
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 06/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 06/28/2019 23:24:14

Diagnostics Check	OK	23:21	
Air Blank	0.000	23:21	
Control Test	0.080	23:22	
Air Blank	0.000	23:22	
Subject Sample	# 1	23:22	Subject Test Refused
Breath Volume	0.000	23:22	
Air Blank	0.000	23:23	
Control Test	0.080	23:23	
Air Blank	0.000	23:24	
Diagnostics Check	OK	23:24	

Data Download Date/Time: 07/25/2019 05:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAYMOND,FRANK,W
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: DAURIA,ANTHONY
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 06/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 05:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:48

SUBJECT'S LAST NAME: MCGORRY

DATE/TIME OF TEST: 06/29/2019 06:16:16

Diagnostics Check	OK	06:13	
Air Blank	0.000	06:13	
Control Test	0.080	06:14	
Air Blank	0.000	06:14	
Subject Sample	# 1	06:14	Subject Test Refused
Breath Volume	0.000	06:14	
Air Blank	0.000	06:15	
Control Test	0.079	06:15	
Air Blank	0.000	06:16	
Diagnostics Check	OK	06:16	

Data Download Date/Time: 07/25/2019 05:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHILE,C,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: BARCARO,B
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 06/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 17:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: JAMISON

DATE/TIME OF TEST: 06/07/2019 18:33:57

Diagnostics Check	OK	18:31	
Air Blank	0.000	18:31	
Control Test	0.082	18:31	
Air Blank	0.000	18:32	
Subject Sample	# 1	18:32	Subject Test Refused
Breath Volume	0.000	18:32	
Air Blank	0.000	18:33	
Control Test	0.082	18:33	
Air Blank	0.000	18:33	
Diagnostics Check	OK	18:33	

Data Download Date/Time: 06/10/2019 19:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEARSE,ANGIE,D
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: BRASHEARS,ERIC
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 06/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: GLOVER

DATE/TIME OF TEST: 06/08/2019 01:51:39

Diagnostics Check	OK	01:44
Air Blank	0.000	01:45
Control Test	0.082	01:45
Air Blank	0.000	01:46
Subject Sample # 1	0.000	01:47
Breath Volume	1.558	01:47
Air Blank	0.000	01:47
Air Blank	0.000	01:49
Subject Sample # 2	0.000	01:50
Breath Volume	1.859	01:50
Air Blank	0.000	01:50
Control Test	0.083	01:51
Air Blank	0.000	01:51
Diagnostics Check	OK	01:51

Data Download Date/Time: 06/10/2019 19:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANDY,KEVIN,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: DONSIFF,MATTHEW
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 06/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: DA SILVA

DATE/TIME OF TEST: 06/09/2019 01:56:54

Diagnostics Check	OK	01:50
Air Blank	0.000	01:50
Control Test	0.083	01:50
Air Blank	0.000	01:51
Subject Sample # 1	0.191	01:52
Breath Volume	1.781	01:52
Air Blank	0.000	01:52
Air Blank	0.000	01:54
Subject Sample # 2	0.184	01:55
Breath Volume	2.128	01:55
Air Blank	0.000	01:55
Control Test	0.083	01:56
Air Blank	0.000	01:56
Diagnostics Check	OK	01:56

Data Download Date/Time: 06/10/2019 19:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHILE,C,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: EGGERS,R
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 06/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 19:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:52

SUBJECT'S LAST NAME: HIGGINS

DATE/TIME OF TEST: 06/11/2019 20:22:46

Diagnostics Check	OK	20:16
Air Blank	0.000	20:16
Control Test	0.083	20:16
Air Blank	0.000	20:17
Subject Sample # 1	0.112	20:18
Breath Volume	1.988	20:18
Air Blank	0.000	20:18
Air Blank	0.000	20:20
Subject Sample # 2	0.120	20:21
Breath Volume	1.984	20:21
Air Blank	0.000	20:21
Control Test	0.083	20:22
Air Blank	0.000	20:22
Diagnostics Check	OK	20:22

Data Download Date/Time: 07/25/2019 05:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLIDAY,LATONYA,A
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: FLORES,DAVID
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 06/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: REMLING

DATE/TIME OF TEST: 06/12/2019 00:29:02

Diagnostics Check	OK	00:22
Air Blank	0.000	00:22
Control Test	0.083	00:23
Air Blank	0.000	00:23
Subject Sample # 1	0.000	00:24
Breath Volume	1.671	00:24
Air Blank	0.000	00:24
Air Blank	0.000	00:26
Subject Sample # 2	0.000	00:27
Breath Volume	1.464	00:27
Air Blank	0.000	00:28
Control Test	0.083	00:28
Air Blank	0.000	00:28
Diagnostics Check	OK	00:29

Data Download Date/Time: 07/25/2019 05:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLIDAY,LATONYA,A
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: SHARPETA,C
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 06/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: PORTZ

DATE/TIME OF TEST: 06/12/2019 03:03:34

Diagnostics Check	OK	02:57
Air Blank	0.000	02:57
Control Test	0.082	02:57
Air Blank	0.000	02:58
Subject Sample # 1	0.176	02:59
Breath Volume	2.417	02:59
Air Blank	0.000	02:59
Air Blank	0.000	03:01
Subject Sample # 2	0.157	03:02
Breath Volume	1.808	03:02
Air Blank	0.000	03:02
Control Test	0.082	03:02
Air Blank	0.000	03:03
Diagnostics Check	OK	03:03

Data Download Date/Time: 07/25/2019 05:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANDY,KEVIN,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: SANTANA,ROMEO
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 06/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: VILLANO

DATE/TIME OF TEST: 06/13/2019 00:22:01

Diagnostics Check	OK	00:15
Air Blank	0.000	00:16
Control Test	0.083	00:16
Air Blank	0.000	00:16
Subject Sample # 1	0.147	00:17
Breath Volume	3.273	00:17
Air Blank	0.000	00:18
Air Blank	0.000	00:19
Subject Sample # 2	0.140	00:20
Breath Volume	3.710	00:20
Air Blank	0.000	00:21
Control Test	0.081	00:21
Air Blank	0.000	00:21
Diagnostics Check	OK	00:22

Data Download Date/Time: 07/25/2019 05:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,TERRENCE,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: JAWORSKI,KEVIN
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 06/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 09:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:02

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 06/14/2019 10:33:08

Diagnostics Check	OK	10:25
Air Blank	0.000	10:25
Control Test	0.082	10:26
Air Blank	0.000	10:26
Subject Sample # 1	0.025	10:28
Breath Volume	1.937	10:28
Air Blank	0.000	10:29
Air Blank	0.000	10:30
Subject Sample # 2	0.023	10:31
Breath Volume	1.671	10:31
Air Blank	0.000	10:32
Control Test	0.082	10:32
Air Blank	0.000	10:33
Diagnostics Check	OK	10:33

Data Download Date/Time: 07/25/2019 05:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLIDAY,LATONYA,A
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: HATTON,N
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 06/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: AUSBY

DATE/TIME OF TEST: 06/20/2019 00:33:00

Diagnostics Check	OK	00:30	
Air Blank	0.000	00:30	
Control Test	0.082	00:30	
Air Blank	0.000	00:31	
Subject Sample	# 1	00:31	Subject Test Refused
Breath Volume	0.000	00:31	
Air Blank	0.000	00:32	
Control Test	0.083	00:32	
Air Blank	0.000	00:32	
Diagnostics Check	OK	00:33	

Data Download Date/Time: 07/25/2019 05:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANDY,KEVIN,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MACHIAS,ERNESTO
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 06/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: JULIANO

DATE/TIME OF TEST: 06/20/2019 01:02:09

Diagnostics Check		OK	00:55
Air Blank		0.000	00:55
Control Test		0.083	00:56
Air Blank		0.000	00:56
Subject Sample	# 1	0.203	00:57
Breath Volume		2.027	00:57
Air Blank		0.000	00:58
Air Blank		0.000	01:00
Subject Sample	# 2	0.200	01:00
Breath Volume		1.816	01:00
Air Blank		0.000	01:01
Control Test		0.082	01:01
Air Blank		0.000	01:02
Diagnostics Check		OK	01:02

Data Download Date/Time: 07/25/2019 05:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,DEVIN,C
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: SHARPETA,CHAD
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 06/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 03:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: DENTON

DATE/TIME OF TEST: 06/20/2019 04:42:15

Diagnostics Check	OK	04:39	
Air Blank	0.000	04:39	
Control Test	0.082	04:40	
Air Blank	0.000	04:40	
Subject Sample	# 1	04:40	Subject Test Refused
Breath Volume	0.000	04:40	
Air Blank	0.000	04:41	
Control Test	0.083	04:41	
Air Blank	0.000	04:42	
Diagnostics Check	OK	04:42	

Data Download Date/Time: 07/25/2019 05:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEARSE,ANGIE,D
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,DENNIS
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 06/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: MEADOWD

DATE/TIME OF TEST: 06/23/2019 01:55:41

Diagnostics Check	OK	01:49
Air Blank	0.000	01:50
Control Test	0.082	01:50
Air Blank	0.000	01:51
Subject Sample # 1	0.000	01:51
Breath Volume	1.441	01:51
Air Blank	0.000	01:52
Air Blank	0.000	01:54
Subject Sample # 2	0.000	01:54
Breath Volume	1.652	01:54
Air Blank	0.000	01:54
Control Test	0.081	01:55
Air Blank	0.000	01:55
Diagnostics Check	OK	01:55

Data Download Date/Time: 07/25/2019 05:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLUMB,GREGORY,T
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: MARTY,KELSEA
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 06/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: ELLIS

DATE/TIME OF TEST: 06/23/2019 04:16:23

Diagnostics Check	OK	04:11	
Air Blank	0.000	04:12	
Control Test	0.082	04:12	
Air Blank	0.000	04:13	
Subject Sample	# 1	04:15	Subject Test Refused
Breath Volume	0.000	04:15	
Air Blank	0.000	04:15	
Control Test	0.083	04:15	
Air Blank	0.000	04:16	
Diagnostics Check	OK	04:16	

Data Download Date/Time: 07/25/2019 05:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,JOHN,P
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: BAKER,DUSTIN
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 05/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: HARDYMAN

DATE/TIME OF TEST: 06/22/2019 00:59:30

Diagnostics Check	OK	00:52
Air Blank	0.000	00:52
Control Test	0.080	00:53
Air Blank	0.000	00:53
Subject Sample # 1	0.184	00:54
Breath Volume	2.757	00:54
Air Blank	0.000	00:55
Air Blank	0.000	00:57
Subject Sample # 2	0.187	00:58
Breath Volume	2.394	00:58
Air Blank	0.000	00:58
Control Test	0.080	00:58
Air Blank	0.000	00:59
Diagnostics Check	OK	00:59

Data Download Date/Time: 06/27/2019 08:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASSEY,DARRELL,L
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: LUTHER,RONNIE
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: PROB/PAROL
ARREST TIME: 19:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: CORTEZ

DATE/TIME OF TEST: 06/30/2019 19:52:24

Diagnostics Check	OK	19:43
Air Blank	0.000	19:43
Control Test	0.080	19:43
Air Blank	0.000	19:44
Subject Sample # 1	0.223	19:47
Breath Volume	1.226	19:47
Air Blank	0.000	19:48
Air Blank	0.000	19:49
Subject Sample # 2	0.214	19:50
Breath Volume	1.324	19:50
Air Blank	0.000	19:51
Control Test	0.080	19:51
Air Blank	0.000	19:52
Diagnostics Check	OK	19:52

Data Download Date/Time: 07/20/2019 16:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WATSON,D,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: WATSON,D
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 06/02/2019 01:53:30

Diagnostics Check	OK	01:50	
Air Blank	0.000	01:51	
Control Test	0.080	01:51	
Air Blank	0.000	01:51	
Subject Sample	# 1	01:52	Subject Test Refused
Breath Volume	0.000	01:52	
Air Blank	0.000	01:52	
Control Test	0.080	01:52	
Air Blank	0.000	01:53	
Diagnostics Check	OK	01:53	

Data Download Date/Time: 06/26/2019 08:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLE,CALVIN,K
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: CARDWELL,CONNER
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: FRENIERE

DATE/TIME OF TEST: 06/02/2019 02:53:23

Diagnostics Check	OK	02:50	
Air Blank	0.000	02:51	
Control Test	0.081	02:51	
Air Blank	0.000	02:51	
Subject Sample	# 1	02:52	Subject Test Refused
Breath Volume	0.000	02:52	
Air Blank	0.000	02:52	
Control Test	0.079	02:52	
Air Blank	0.000	02:53	
Diagnostics Check	OK	02:53	

Data Download Date/Time: 06/26/2019 08:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDEZ,DEPUTY,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: CHAPMAN,OFFICER
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 05/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: BUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 06/02/2019 04:47:31

Diagnostics Check	OK	04:44	
Air Blank	0.000	04:44	
Control Test	0.080	04:45	
Air Blank	0.000	04:45	
Subject Sample	# 1	04:46	Subject Test Refused
Breath Volume	0.000	04:46	
Air Blank	0.000	04:46	
Control Test	0.080	04:46	
Air Blank	0.000	04:47	
Diagnostics Check	OK	04:47	

Data Download Date/Time: 06/26/2019 08:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERSON,STEFANIE,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PETERSON,STEFANIE
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 05:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:48

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 06/02/2019 06:21:42

Diagnostics Check		OK	06:11
Air Blank		0.000	06:12
Control Test		0.079	06:12
Air Blank		0.000	06:12
Subject Sample	# 1	0.250	06:14
Breath Volume		1.941	06:14
Air Blank		0.000	06:15
Air Blank		0.000	06:16
Subject Sample	# 2	0.225	06:17
Breath Volume		2.566	06:17
Air Blank		0.000	06:17
Air Blank		0.000	06:19
Subject Sample	# 3	0.223	06:20
Breath Volume		1.925	06:20
Air Blank		0.000	06:20
Control Test		0.078	06:21
Air Blank		0.000	06:21
Diagnostics Check		OK	06:21

Data Download Date/Time: 06/26/2019 08:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDEZ,DEPUTY,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MILLER,DEPUTY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:11

SUBJECT'S LAST NAME: VERGARA LOPEZ

DATE/TIME OF TEST: 06/02/2019 22:43:21

Diagnostics Check	OK	22:36
Air Blank	0.000	22:36
Control Test	0.080	22:37
Air Blank	0.000	22:37
Subject Sample # 1	0.122	22:38
Breath Volume	1.917	22:38
Air Blank	0.000	22:39
Air Blank	0.000	22:41
Subject Sample # 2	0.118	22:41
Breath Volume	2.449	22:41
Air Blank	0.000	22:42
Control Test	0.079	22:42
Air Blank	0.000	22:43
Diagnostics Check	OK	22:43

Data Download Date/Time: 06/26/2019 08:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERSON,STEFANIE,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MARTIN,SHAUN
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 06/02/2019 23:53:22

Diagnostics Check	OK	23:43
Air Blank	0.000	23:44
Control Test	0.080	23:44
Air Blank	0.000	23:45
Subject Sample # 1	0.130	23:45
Breath Volume	1.339	23:45
Air Blank	0.000	23:46
Air Blank	0.000	23:48
Subject Sample # 2	0.103	23:48
Breath Volume	1.437	23:48
Air Blank	0.000	23:49
Air Blank	0.000	23:51
Subject Sample # 3	0.107	23:51
Breath Volume	1.906	23:51
Air Blank	0.000	23:52
Control Test	0.080	23:52
Air Blank	0.000	23:53
Diagnostics Check	OK	23:53

Data Download Date/Time: 06/26/2019 08:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,J,S
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BARR,J
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: FERRARA

DATE/TIME OF TEST: 06/03/2019 22:53:45

Diagnostics Check	OK	22:47
Air Blank	0.000	22:47
Control Test	0.080	22:47
Air Blank	0.000	22:48
Subject Sample # 1	0.209	22:49
Breath Volume	2.636	22:49
Air Blank	0.000	22:49
Air Blank	0.000	22:51
Subject Sample # 2	0.202	22:52
Breath Volume	2.359	22:52
Air Blank	0.000	22:52
Control Test	0.079	22:53
Air Blank	0.000	22:53
Diagnostics Check	OK	22:53

Data Download Date/Time: 06/26/2019 08:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,NICKKI,N
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MONTANEZ,VICTOR
ARRESTING OFFICER AGENCY: Venice PD
LAST AGENCY INSPECTION: 05/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:51

SUBJECT'S LAST NAME: BROOKER

DATE/TIME OF TEST: 06/04/2019 21:31:27

Diagnostics Check	OK	21:24
Air Blank	0.000	21:25
Control Test	0.080	21:25
Air Blank	0.000	21:26
Subject Sample # 1	0.110	21:26
Breath Volume	1.527	21:26
Air Blank	0.000	21:27
Air Blank	0.000	21:29
Subject Sample # 2	0.124	21:29
Breath Volume	2.011	21:29
Air Blank	0.000	21:30
Control Test	0.080	21:30
Air Blank	0.000	21:31
Diagnostics Check	OK	21:31

Data Download Date/Time: 06/26/2019 08:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDEZ,DEPUTY,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MARTIN,DEPUTY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: HENSE

DATE/TIME OF TEST: 06/07/2019 01:36:20

Diagnostics Check		OK	01:28
Air Blank		0.000	01:29
Control Test		0.080	01:29
Air Blank		0.000	01:30
Subject Sample	# 1	0.098	01:31
Breath Volume		1.460	01:31
Air Blank		0.000	01:31
Air Blank		0.000	01:33
Subject Sample	# 2	0.102	01:34
Breath Volume		1.617	01:34
Air Blank		0.000	01:35
Control Test		0.081	01:35
Air Blank		0.000	01:36
Diagnostics Check		OK	01:36

Data Download Date/Time: 06/26/2019 08:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAMORS,JASON,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: TRASKA,L
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: NOYES

DATE/TIME OF TEST: 06/07/2019 02:45:01

Diagnostics Check	OK	02:38
Air Blank	0.000	02:39
Control Test	0.080	02:39
Air Blank	0.000	02:40
Subject Sample # 1	0.149	02:40
Breath Volume	2.027	02:40
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.148	02:43
Breath Volume	1.851	02:43
Air Blank	0.000	02:44
Control Test	0.080	02:44
Air Blank	0.000	02:44
Diagnostics Check	OK	02:45

Data Download Date/Time: 06/26/2019 08:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACEVEDO VALDES,CRISTIAN,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ACEVEDO VALDES,CRISTIAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 06:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:33

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 06/08/2019 08:07:27

Diagnostics Check		OK	08:01
Air Blank		0.000	08:01
Control Test		0.080	08:02
Air Blank		0.000	08:02
Subject Sample	# 1	0.064	08:03
Breath Volume		2.707	08:03
Air Blank		0.000	08:03
Air Blank		0.000	08:05
Subject Sample	# 2	0.068	08:05
Breath Volume		2.816	08:05
Air Blank		0.000	08:06
Control Test		0.080	08:06
Air Blank		0.000	08:07
Diagnostics Check		OK	08:07

Data Download Date/Time: 06/26/2019 08:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FELTOVIC,DEP,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FELTOVIC,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 11:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:16

SUBJECT'S LAST NAME: FOTHERGILL

DATE/TIME OF TEST: 06/08/2019 12:48:33

Diagnostics Check	OK	12:41
Air Blank	0.000	12:42
Control Test	0.081	12:42
Air Blank	0.000	12:43
Subject Sample # 1	0.211	12:43
Breath Volume	1.992	12:43
Air Blank	0.000	12:44
Air Blank	0.000	12:45
Subject Sample # 2	0.224	12:46
Breath Volume	2.085	12:46
Air Blank	0.000	12:47
Control Test	0.080	12:47
Air Blank	0.000	12:48
Diagnostics Check	OK	12:48

Data Download Date/Time: 06/26/2019 08:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACOSTA-BURGOS,KASSANDRA,J
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ACOSTA-
BURGOS,KASSANDRA
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:07
GENDER: Male

OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: CALIRI

DATE/TIME OF TEST: 06/09/2019 00:39:23

Diagnostics Check	OK	00:36	
Air Blank	0.000	00:37	
Control Test	0.080	00:37	
Air Blank	0.000	00:37	
Subject Sample	# 1	00:38	Subject Test Refused
Breath Volume	0.000	00:38	
Air Blank	0.000	00:38	
Control Test	0.079	00:38	
Air Blank	0.000	00:39	
Diagnostics Check	OK	00:39	

Data Download Date/Time: 06/26/2019 08:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDEZ,DEPUTY,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MARTIN,DEPUTY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:17

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 06/11/2019 21:49:34

Diagnostics Check	OK	21:41
Air Blank	0.000	21:42
Control Test	0.080	21:42
Air Blank	0.000	21:43
Subject Sample # 1	0.158	21:44
Breath Volume	1.855	21:44
Air Blank	0.000	21:45
Air Blank	0.000	21:47
Subject Sample # 2	0.159	21:47
Breath Volume	1.343	21:47
Air Blank	0.000	21:48
Control Test	0.077	21:48
Air Blank	0.000	21:49
Diagnostics Check	OK	21:49

Data Download Date/Time: 06/26/2019 08:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUMANN,DEP,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SCHUMANN,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: ETCHEVERRY

DATE/TIME OF TEST: 06/13/2019 00:52:23

Diagnostics Check	OK	00:46
Air Blank	0.000	00:46
Control Test	0.079	00:46
Air Blank	0.000	00:47
Subject Sample # 1	0.171	00:47
Breath Volume	2.742	00:47
Air Blank	0.000	00:48
Air Blank	0.000	00:50
Subject Sample # 2	0.167	00:50
Breath Volume	1.812	00:50
Air Blank	0.000	00:51
Control Test	0.079	00:51
Air Blank	0.000	00:52
Diagnostics Check	OK	00:52

Data Download Date/Time: 06/26/2019 08:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLINT,SERGEANT,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FLINT,SERGEANT
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:31

SUBJECT'S LAST NAME: LANSLEY

DATE/TIME OF TEST: 06/13/2019 20:55:08

Diagnostics Check	OK	20:52	
Air Blank	0.000	20:52	
Control Test	0.080	20:52	
Air Blank	0.000	20:53	
Subject Sample	# 1	20:53	Subject Test Refused
Breath Volume	0.000	20:53	
Air Blank	0.000	20:54	
Control Test	0.080	20:54	
Air Blank	0.000	20:55	
Diagnostics Check	OK	20:55	

Data Download Date/Time: 06/26/2019 08:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VERMILLION,B,
BREATH TEST OPERATOR AGENCY: Sarasota PD
ARRESTING OFFICER: VERMILLION,B
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 05/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 11:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:51

SUBJECT'S LAST NAME: FORSON

DATE/TIME OF TEST: 06/14/2019 12:20:46

Diagnostics Check	OK	12:14
Air Blank	0.000	12:15
Control Test	0.080	12:15
Air Blank	0.000	12:16
Subject Sample # 1	0.000	12:16
Breath Volume	3.273	12:16
Air Blank	0.000	12:16
Air Blank	0.000	12:18
Subject Sample # 2	0.000	12:19
Breath Volume	3.371	12:19
Air Blank	0.000	12:19
Control Test	0.079	12:20
Air Blank	0.000	12:20
Diagnostics Check	OK	12:20

Data Download Date/Time: 06/26/2019 08:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLINT,SERGEANT,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FLINT,SERGEANT
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: DELAHANTY

DATE/TIME OF TEST: 06/14/2019 21:46:56

Diagnostics Check		OK	21:35	
Air Blank		0.000	21:35	
Control Test		0.080	21:36	
Air Blank		0.000	21:36	
Subject Sample	# 1	0.054	21:39	Volume Not Met
Breath Volume		0.078	21:39	
Air Blank		0.000	21:40	
Air Blank		0.000	21:42	
Subject Sample	# 2	0.063	21:45	Volume Not Met
Breath Volume		0.121	21:45	
Air Blank		0.000	21:46	
Control Test		0.078	21:46	
Air Blank		0.000	21:46	
Diagnostics Check		OK	21:46	

Data Download Date/Time: 06/26/2019 08:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLINT,SERGEANT,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FLINT,SERGEANT
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: DELAHANTY

DATE/TIME OF TEST: 06/14/2019 22:02:28

Diagnostics Check		OK	21:48	
Air Blank		0.000	21:49	
Control Test		0.076	21:49	
Air Blank		0.000	21:50	
Subject Sample	# 1	0.055	21:53	Volume Not Met
Breath Volume		0.285	21:53	
Air Blank		0.000	21:54	
Air Blank		0.000	21:55	
Subject Sample	# 2	0.061	21:56	
Breath Volume		2.375	21:56	
Air Blank		0.000	21:57	
Air Blank		0.000	21:59	
Subject Sample	# 3	0.064	22:00	
Breath Volume		1.964	22:00	
Air Blank		0.000	22:01	
Control Test		0.079	22:01	
Air Blank		0.000	22:02	
Diagnostics Check		OK	22:02	

Data Download Date/Time: 06/26/2019 08:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLINT,SERGEANT,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BELL,DEPUTY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 06/14/2019 23:53:05

Diagnostics Check	OK	23:44
Air Blank	0.000	23:45
Control Test	0.079	23:45
Air Blank	0.000	23:46
Subject Sample # 1	0.126	23:48
Breath Volume	2.074	23:48
Air Blank	0.000	23:49
Air Blank	0.000	23:50
Subject Sample # 2	0.130	23:51
Breath Volume	1.847	23:51
Air Blank	0.000	23:52
Control Test	0.079	23:52
Air Blank	0.000	23:53
Diagnostics Check	OK	23:53

Data Download Date/Time: 06/26/2019 08:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERSON,STEFANIE,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PETERSON,STEFANIE
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 06/15/2019 04:31:52

Diagnostics Check	OK	04:25
Air Blank	0.000	04:25
Control Test	0.080	04:25
Air Blank	0.000	04:26
Subject Sample # 1	0.090	04:27
Breath Volume	1.402	04:27
Air Blank	0.000	04:28
Air Blank	0.000	04:29
Subject Sample # 2	0.074	04:30
Breath Volume	1.269	04:30
Air Blank	0.000	04:30
Control Test	0.078	04:31
Air Blank	0.000	04:31
Diagnostics Check	OK	04:31

Data Download Date/Time: 06/26/2019 08:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/28/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:14

SUBJECT'S LAST NAME: WOLOSZ

DATE/TIME OF TEST: 06/16/2019 21:43:12

Diagnostics Check	OK	21:36
Air Blank	0.000	21:36
Control Test	0.081	21:36
Air Blank	0.000	21:37
Subject Sample # 1	0.139	21:38
Breath Volume	2.378	21:38
Air Blank	0.000	21:38
Air Blank	0.000	21:40
Subject Sample # 2	0.146	21:41
Breath Volume	3.488	21:41
Air Blank	0.000	21:42
Control Test	0.079	21:42
Air Blank	0.000	21:43
Diagnostics Check	OK	21:43

Data Download Date/Time: 06/26/2019 08:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FELTOVIC,DEP,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FINNEGAN,E
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 05/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 11:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:10

SUBJECT'S LAST NAME: RICHARD

DATE/TIME OF TEST: 06/18/2019 13:38:38

Diagnostics Check	OK	13:32
Air Blank	0.000	13:33
Control Test	0.081	13:33
Air Blank	0.000	13:34
Subject Sample # 1	0.000	13:34
Breath Volume	2.050	13:34
Air Blank	0.000	13:34
Air Blank	0.000	13:36
Subject Sample # 2	0.000	13:37
Breath Volume	1.742	13:37
Air Blank	0.000	13:37
Control Test	0.079	13:38
Air Blank	0.000	13:38
Diagnostics Check	OK	13:38

Data Download Date/Time: 06/26/2019 08:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,N,N
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: JEWELL,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: OTHER
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 06/19/2019 02:03:12

Diagnostics Check	OK	01:56
Air Blank	0.000	01:57
Control Test	0.080	01:57
Air Blank	0.000	01:58
Subject Sample # 1	0.175	01:58
Breath Volume	2.363	01:58
Air Blank	0.000	01:59
Air Blank	0.000	02:01
Subject Sample # 2	0.167	02:01
Breath Volume	1.324	02:01
Air Blank	0.000	02:02
Control Test	0.079	02:02
Air Blank	0.000	02:03
Diagnostics Check	OK	02:03

Data Download Date/Time: 06/26/2019 08:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSTYAL,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KOSTYAL,MATTHEW
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: WARWICK

DATE/TIME OF TEST: 06/20/2019 03:35:42

Diagnostics Check	OK	03:28
Air Blank	0.000	03:29
Control Test	0.081	03:29
Air Blank	0.000	03:30
Subject Sample # 1	0.063	03:30
Breath Volume	1.636	03:30
Air Blank	0.000	03:31
Air Blank	0.000	03:33
Subject Sample # 2	0.062	03:34
Breath Volume	1.429	03:34
Air Blank	0.000	03:34
Control Test	0.080	03:35
Air Blank	0.000	03:35
Diagnostics Check	OK	03:35

Data Download Date/Time: 06/26/2019 08:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUEHLER,DEP,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BUEHLER,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:48

SUBJECT'S LAST NAME: HENNIGE

DATE/TIME OF TEST: 06/20/2019 21:14:58

Diagnostics Check	OK	21:08
Air Blank	0.000	21:09
Control Test	0.078	21:09
Air Blank	0.000	21:09
Subject Sample # 1	0.173	21:10
Breath Volume	3.843	21:10
Air Blank	0.000	21:11
Air Blank	0.000	21:12
Subject Sample # 2	0.171	21:13
Breath Volume	3.968	21:13
Air Blank	0.000	21:14
Control Test	0.077	21:14
Air Blank	0.000	21:14
Diagnostics Check	OK	21:14

Data Download Date/Time: 06/26/2019 08:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACOSTA-BURGOS,KASSANDRA,J
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ACOSTA-
BURGOS,KASSANDRA
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male

OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: LOPEZ LOPEZ

DATE/TIME OF TEST: 06/23/2019 04:05:31

Diagnostics Check	OK	03:57	
Air Blank	0.000	03:57	
Control Test	0.081	03:57	
Air Blank	0.000	03:58	
Subject Sample	# 1 0.146	04:01	Volume Not Met
Breath Volume	1.046	04:01	
Air Blank	0.000	04:02	
Air Blank	0.000	04:03	
Subject Sample	# 2	04:04	RFI Detect
Breath Volume	0.000	04:04	
Air Blank	0.000	04:04	
Control Test	0.077	04:04	
Air Blank	0.000	04:05	
Diagnostics Check	OK	04:05	

Data Download Date/Time: 06/26/2019 08:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACOSTA-BURGOS,KASSANDRA,J
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ACOSTA-
BURGOS,KASSANDRA
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male

OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: LOPEZ LOPEZ

DATE/TIME OF TEST: 06/23/2019 04:15:59

Diagnostics Check	OK	04:07
Air Blank	0.000	04:07
Control Test	0.079	04:08
Air Blank	0.000	04:08
Subject Sample # 1	0.148	04:10
Breath Volume	2.066	04:10
Air Blank	0.000	04:11
Air Blank	0.000	04:13
Subject Sample # 2	0.134	04:14
Breath Volume	1.210	04:14
Air Blank	0.000	04:15
Control Test	0.076	04:15
Air Blank	0.000	04:15
Diagnostics Check	OK	04:15

Data Download Date/Time: 06/26/2019 08:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDEZ,DEPUTY,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MILLER,DEPUTY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 06/24/2019 22:55:02

Diagnostics Check	OK	22:45
Air Blank	0.000	22:45
Control Test	0.080	22:46
Air Blank	0.000	22:46
Subject Sample # 1	0.215	22:47
Breath Volume	2.113	22:47
Air Blank	0.000	22:48
Air Blank	0.000	22:49
Subject Sample # 2	0.236	22:50
Breath Volume	2.449	22:50
Air Blank	0.000	22:51
Air Blank	0.000	22:52
Subject Sample # 3	0.211	22:53
Breath Volume	1.648	22:53
Air Blank	0.000	22:54
Control Test	0.079	22:54
Air Blank	0.000	22:54
Diagnostics Check	OK	22:55

Data Download Date/Time: 06/26/2019 08:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAMORS,JASON,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: TURNER,B
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: KLIJEWSKI

DATE/TIME OF TEST: 06/24/2019 23:47:18

Diagnostics Check	OK	23:39
Air Blank	0.000	23:40
Control Test	0.079	23:40
Air Blank	0.000	23:41
Subject Sample # 1	0.146	23:42
Breath Volume	1.574	23:42
Air Blank	0.000	23:43
Air Blank	0.000	23:44
Subject Sample # 2	0.150	23:45
Breath Volume	1.875	23:45
Air Blank	0.000	23:46
Control Test	0.078	23:46
Air Blank	0.000	23:47
Diagnostics Check	OK	23:47

Data Download Date/Time: 06/26/2019 08:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WATSON,D,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: WATSON,D
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: CHAPMAN

DATE/TIME OF TEST: 06/25/2019 02:00:58

Diagnostics Check	OK	01:54
Air Blank	0.000	01:55
Control Test	0.079	01:55
Air Blank	0.000	01:55
Subject Sample # 1	0.147	01:56
Breath Volume	3.445	01:56
Air Blank	0.000	01:57
Air Blank	0.000	01:58
Subject Sample # 2	0.140	01:59
Breath Volume	3.386	01:59
Air Blank	0.000	02:00
Control Test	0.078	02:00
Air Blank	0.000	02:00
Diagnostics Check	OK	02:00

Data Download Date/Time: 06/26/2019 08:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUMANN,DEP,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SCHUMANN,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: HAAS

DATE/TIME OF TEST: 06/25/2019 23:27:08

Diagnostics Check	OK	23:16
Air Blank	0.000	23:17
Control Test	0.076	23:17
Air Blank	0.000	23:18
Subject Sample # 1	0.244	23:20
Breath Volume	1.562	23:20
Air Blank	0.000	23:21
Air Blank	0.000	23:22
Subject Sample # 2	0.226	23:25
Breath Volume	1.949	23:25
Air Blank	0.000	23:26
Control Test	0.076	23:26
Air Blank	0.000	23:27
Diagnostics Check	OK	23:27

Data Download Date/Time: 06/26/2019 08:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANTWELL,J,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: CANTWELL,J
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: LUND

DATE/TIME OF TEST: 06/26/2019 00:50:12

Diagnostics Check	OK	00:43
Air Blank	0.000	00:43
Control Test	0.079	00:44
Air Blank	0.000	00:44
Subject Sample # 1	0.190	00:45
Breath Volume	1.691	00:45
Air Blank	0.000	00:46
Air Blank	0.000	00:47
Subject Sample # 2	0.203	00:48
Breath Volume	1.929	00:48
Air Blank	0.000	00:49
Control Test	0.079	00:49
Air Blank	0.000	00:50
Diagnostics Check	OK	00:50

Data Download Date/Time: 06/26/2019 08:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WATSON,D,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: WATSON,D
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: KRAEMER

DATE/TIME OF TEST: 06/26/2019 02:14:41

Diagnostics Check	OK	02:08
Air Blank	0.000	02:08
Control Test	0.079	02:08
Air Blank	0.000	02:09
Subject Sample # 1	0.177	02:10
Breath Volume	1.589	02:10
Air Blank	0.000	02:10
Air Blank	0.000	02:12
Subject Sample # 2	0.176	02:13
Breath Volume	1.128	02:13
Air Blank	0.000	02:13
Control Test	0.078	02:14
Air Blank	0.000	02:14
Diagnostics Check	OK	02:14

Data Download Date/Time: 06/26/2019 08:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER, TYLER, J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FOWLER, T
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: GOZDAL

DATE/TIME OF TEST: 06/26/2019 02:50:54

Diagnostics Check	OK	02:44
Air Blank	0.000	02:45
Control Test	0.079	02:45
Air Blank	0.000	02:45
Subject Sample # 1	0.139	02:46
Breath Volume	3.406	02:46
Air Blank	0.000	02:47
Air Blank	0.000	02:48
Subject Sample # 2	0.135	02:49
Breath Volume	3.664	02:49
Air Blank	0.000	02:49
Control Test	0.075	02:50
Air Blank	0.000	02:50
Diagnostics Check	OK	02:50

Data Download Date/Time: 06/26/2019 08:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDEZ,DEPUTY,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FOLADVANDI,OFFICER
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 05/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: SWEENEY

DATE/TIME OF TEST: 06/26/2019 03:15:36

Diagnostics Check	OK	03:06
Air Blank	0.000	03:07
Control Test	0.078	03:07
Air Blank	0.000	03:08
Subject Sample # 1	0.103	03:10
Breath Volume	1.476	03:10
Air Blank	0.000	03:11
Air Blank	0.000	03:13
Subject Sample # 2	0.104	03:14
Breath Volume	1.691	03:14
Air Blank	0.000	03:14
Control Test	0.077	03:15
Air Blank	0.000	03:15
Diagnostics Check	OK	03:15

Data Download Date/Time: 06/26/2019 08:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,DAVID,A
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: PEREZ,DAVID
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 05/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 06/20/2019 02:38:31

Diagnostics Check		OK	02:32
Air Blank		0.000	02:33
Control Test		0.080	02:33
Air Blank		0.000	02:33
Subject Sample	# 1	0.166	02:34
Breath Volume		2.199	02:34
Air Blank		0.000	02:34
Air Blank		0.000	02:36
Subject Sample	# 2	0.168	02:36
Breath Volume		2.167	02:36
Air Blank		0.000	02:37
Control Test		0.080	02:37
Air Blank		0.000	02:38
Diagnostics Check		OK	02:38

Data Download Date/Time: 07/01/2019 16:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HACKETT,SEAN,J
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: WILLIAMS,AUSTIN
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 05/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 14:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: DOWDA

DATE/TIME OF TEST: 06/21/2019 15:36:00

Diagnostics Check	OK	15:32	
Air Blank	0.000	15:33	
Control Test	0.080	15:33	
Air Blank	0.000	15:34	
Subject Sample	# 1	15:34	Subject Test Refused
Breath Volume	0.000	15:34	
Air Blank	0.000	15:35	
Control Test	0.080	15:35	
Air Blank	0.000	15:35	
Diagnostics Check	OK	15:36	

Data Download Date/Time: 07/01/2019 16:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,DAVID,A
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: PINSON,NICHOLAS
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 05/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: BATTIN

DATE/TIME OF TEST: 06/30/2019 01:40:38

Diagnostics Check		OK	01:34
Air Blank		0.000	01:34
Control Test		0.081	01:35
Air Blank		0.000	01:35
Subject Sample	# 1	0.244	01:36
Breath Volume		2.609	01:36
Air Blank		0.000	01:36
Air Blank		0.000	01:38
Subject Sample	# 2	0.235	01:39
Breath Volume		2.082	01:39
Air Blank		0.000	01:39
Control Test		0.080	01:40
Air Blank		0.000	01:40
Diagnostics Check		OK	01:40

Data Download Date/Time: 07/01/2019 16:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STROUP,STEVE,A
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: RICE,ANDREW
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 05/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 15:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:07

SUBJECT'S LAST NAME: KIDD

DATE/TIME OF TEST: 06/30/2019 16:47:47

Diagnostics Check	OK	16:37
Air Blank	0.000	16:37
Control Test	0.079	16:37
Air Blank	0.000	16:38
Subject Sample # 1	0.279	16:39
Breath Volume	2.390	16:39
Air Blank	0.000	16:40
Air Blank	0.000	16:41
Subject Sample # 2	0.307	16:43
Breath Volume	2.078	16:43
Air Blank	0.000	16:44
Air Blank	0.000	16:45
Subject Sample # 3	0.304	16:46
Breath Volume	1.488	16:46
Air Blank	0.000	16:46
Control Test	0.076	16:47
Air Blank	0.000	16:47
Diagnostics Check	OK	16:47

Data Download Date/Time: 07/01/2019 16:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUMANN,DEP,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SCHUMANN,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: VU

DATE/TIME OF TEST: 06/29/2019 03:06:43

Diagnostics Check	OK	03:00
Air Blank	0.000	03:00
Control Test	0.079	03:01
Air Blank	0.000	03:01
Subject Sample # 1	0.171	03:02
Breath Volume	2.589	03:02
Air Blank	0.000	03:02
Air Blank	0.000	03:04
Subject Sample # 2	0.159	03:05
Breath Volume	2.890	03:05
Air Blank	0.000	03:05
Control Test	0.078	03:06
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

Data Download Date/Time: 07/30/2019 08:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUMANN,DEP,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: WATSON,D
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:14

SUBJECT'S LAST NAME: JABLONSKI

DATE/TIME OF TEST: 06/29/2019 22:30:51

Diagnostics Check	OK	22:23
Air Blank	0.000	22:24
Control Test	0.080	22:24
Air Blank	0.000	22:24
Subject Sample # 1	0.113	22:26
Breath Volume	3.585	22:26
Air Blank	0.000	22:26
Air Blank	0.000	22:28
Subject Sample # 2	0.118	22:29
Breath Volume	3.957	22:29
Air Blank	0.000	22:29
Control Test	0.078	22:30
Air Blank	0.000	22:30
Diagnostics Check	OK	22:30

Data Download Date/Time: 07/30/2019 08:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUEHLER,DEP,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BUEHLER,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 06/28/2019 03:27:06

Diagnostics Check		OK	03:20
Air Blank		0.000	03:21
Control Test		0.078	03:21
Air Blank		0.000	03:21
Subject Sample	# 1	0.168	03:22
Breath Volume		1.769	03:22
Air Blank		0.000	03:23
Air Blank		0.000	03:25
Subject Sample	# 2	0.170	03:25
Breath Volume		1.613	03:25
Air Blank		0.000	03:26
Control Test		0.078	03:26
Air Blank		0.000	03:27
Diagnostics Check		OK	03:27

Data Download Date/Time: 07/30/2019 07:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER, TYLER, J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FOWLER, TYLER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:23

SUBJECT'S LAST NAME: ROUSES

DATE/TIME OF TEST: 06/28/2019 23:51:48

Diagnostics Check	OK	23:45
Air Blank	0.000	23:45
Control Test	0.078	23:46
Air Blank	0.000	23:46
Subject Sample # 1	0.335	23:47
Breath Volume	2.554	23:47
Air Blank	0.000	23:48
Air Blank	0.000	23:49
Subject Sample # 2	0.320	23:50
Breath Volume	1.992	23:50
Air Blank	0.000	23:50
Control Test	0.077	23:51
Air Blank	0.000	23:51
Diagnostics Check	OK	23:51

Data Download Date/Time: 07/30/2019 07:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER,T,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BUTLER,C
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: BRIERE

DATE/TIME OF TEST: 06/29/2019 00:44:09

Diagnostics Check	OK	00:37
Air Blank	0.000	00:38
Control Test	0.079	00:38
Air Blank	0.000	00:39
Subject Sample # 1	0.178	00:39
Breath Volume	3.785	00:39
Air Blank	0.000	00:40
Air Blank	0.000	00:42
Subject Sample # 2	0.166	00:42
Breath Volume	1.589	00:42
Air Blank	0.000	00:43
Control Test	0.077	00:43
Air Blank	0.000	00:44
Diagnostics Check	OK	00:44

Data Download Date/Time: 07/30/2019 07:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER,T,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FOWLER,T
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: CORDERO

DATE/TIME OF TEST: 06/29/2019 21:41:47

Diagnostics Check	OK	21:35
Air Blank	0.000	21:36
Control Test	0.078	21:36
Air Blank	0.000	21:37
Subject Sample # 1	0.117	21:37
Breath Volume	2.636	21:37
Air Blank	0.000	21:38
Air Blank	0.000	21:40
Subject Sample # 2	0.115	21:40
Breath Volume	2.656	21:40
Air Blank	0.000	21:40
Control Test	0.078	21:41
Air Blank	0.000	21:41
Diagnostics Check	OK	21:41

Data Download Date/Time: 07/30/2019 07:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WATTS,LARESSA,G
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GUYTON,JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 05/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: WATTS

DATE/TIME OF TEST: 06/03/2019 04:13:54

Diagnostics Check	OK	04:13	
Air Blank		04:13	RFI Detect
Air Blank	0.000	04:13	

Data Download Date/Time: 06/26/2019 19:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPINGER,ALEC,R
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GUYTON,JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 05/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: WATTS

DATE/TIME OF TEST: 06/03/2019 04:27:09

Diagnostics Check	OK	04:15	
Air Blank	0.000	04:16	
Control Test	0.078	04:16	
Air Blank	0.000	04:17	
Subject Sample	# 1	04:20	No Sample Provided
Breath Volume	0.000	04:20	
Air Blank	0.000	04:20	
Air Blank	0.000	04:22	
Subject Sample	# 2	04:25	No Sample Provided
Breath Volume	0.000	04:25	
Air Blank	0.000	04:26	
Control Test	0.079	04:26	
Air Blank	0.000	04:27	
Diagnostics Check	OK	04:27	

Data Download Date/Time: 06/26/2019 19:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEIMIG,EMORIE,I
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: TURNER,KENDRICK
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 05/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 06/05/2019 00:33:44

Diagnostics Check	OK	00:27
Air Blank	0.000	00:27
Control Test	0.079	00:28
Air Blank	0.000	00:28
Subject Sample # 1	0.000	00:29
Breath Volume	1.472	00:29
Air Blank	0.000	00:29
Air Blank	0.000	00:31
Subject Sample # 2	0.000	00:32
Breath Volume	3.031	00:32
Air Blank	0.000	00:32
Control Test	0.079	00:33
Air Blank	0.000	00:33
Diagnostics Check	OK	00:33

Data Download Date/Time: 06/26/2019 19:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAGE,DEBRA,J
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: DASHER,CHRIS
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 05/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:49

SUBJECT'S LAST NAME: PETTERSSON

DATE/TIME OF TEST: 06/12/2019 19:21:44

Diagnostics Check	OK	19:15
Air Blank	0.000	19:16
Control Test	0.078	19:16
Air Blank	0.000	19:16
Subject Sample # 1	0.238	19:17
Breath Volume	2.316	19:17
Air Blank	0.000	19:17
Air Blank	0.000	19:19
Subject Sample # 2	0.229	19:20
Breath Volume	2.085	19:20
Air Blank	0.000	19:20
Control Test	0.077	19:21
Air Blank	0.000	19:21
Diagnostics Check	OK	19:21

Data Download Date/Time: 06/26/2019 19:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWAIN,DUSTIN,J
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: TURNER,KENDRICK
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 05/22/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: SWAIN

DATE/TIME OF TEST: 06/14/2019 03:53:36

Diagnostics Check	OK	03:50	
Air Blank	0.000	03:50	
Control Test	0.080	03:51	
Air Blank	0.000	03:51	
Subject Sample	# 1	03:52	Subject Test Refused
Breath Volume	0.000	03:52	
Air Blank	0.000	03:52	
Control Test	0.080	03:53	
Air Blank	0.000	03:53	
Diagnostics Check	OK	03:53	

Data Download Date/Time: 06/26/2019 19:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURWELL,MARK,J
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: TURNER,KENDRICK
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 05/22/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: SWAIN

DATE/TIME OF TEST: 06/14/2019 04:10:49

Diagnostics Check	OK	04:04
Air Blank	0.000	04:05
Control Test	0.080	04:05
Air Blank	0.000	04:05
Subject Sample # 1	0.128	04:06
Breath Volume	2.839	04:06
Air Blank	0.000	04:06
Air Blank	0.000	04:08
Subject Sample # 2	0.129	04:09
Breath Volume	3.218	04:09
Air Blank	0.000	04:09
Control Test	0.080	04:10
Air Blank	0.000	04:10
Diagnostics Check	OK	04:10

Data Download Date/Time: 06/26/2019 19:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAGE,DEBRA,J
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: PAGE,DEBRA
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 05/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 13:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:02

SUBJECT'S LAST NAME: THORNTON

DATE/TIME OF TEST: 06/14/2019 14:38:23

Diagnostics Check	OK	14:30
Air Blank	0.000	14:31
Control Test	0.079	14:31
Air Blank	0.000	14:31
Subject Sample # 1	0.000	14:32
Breath Volume	1.714	14:32
Air Blank	0.000	14:33
Air Blank	0.000	14:35
Subject Sample # 2	0.000	14:37
Breath Volume	2.746	14:37
Air Blank	0.000	14:37
Control Test	0.079	14:37
Air Blank	0.000	14:38
Diagnostics Check	OK	14:38

Data Download Date/Time: 06/26/2019 19:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,KENDRICK,T
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: TURNER,KENDRICK
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 05/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 06/15/2019 02:15:31

Diagnostics Check	OK	02:04	
Air Blank	0.000	02:04	
Control Test	0.079	02:04	
Air Blank	0.000	02:05	
Subject Sample	# 1	02:08	No Sample Provided
Breath Volume	0.000	02:08	
Air Blank	0.000	02:09	
Air Blank	0.000	02:11	
Subject Sample	# 2	02:14	No Sample Provided
Breath Volume	0.000	02:14	
Air Blank	0.000	02:14	
Control Test	0.080	02:14	
Air Blank	0.000	02:15	
Diagnostics Check	OK	02:15	

Data Download Date/Time: 06/26/2019 19:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,CHRISTOPHER,
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GUYTON,JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 05/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: BRYDEN

DATE/TIME OF TEST: 06/21/2019 02:52:44

Diagnostics Check	OK	02:49	
Air Blank	0.000	02:50	
Control Test	0.080	02:50	
Air Blank	0.000	02:51	
Subject Sample	# 1	02:51	Subject Test Refused
Breath Volume	0.000	02:51	
Air Blank	0.000	02:51	
Control Test	0.079	02:52	
Air Blank	0.000	02:52	
Diagnostics Check	OK	02:52	

Data Download Date/Time: 06/26/2019 19:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,CHRISTOPHER,
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GUYTON,JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 06/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: HOWDERSHELT

DATE/TIME OF TEST: 06/24/2019 23:47:59

Diagnostics Check	OK	23:45	
Air Blank	0.000	23:45	
Control Test	0.080	23:45	
Air Blank	0.000	23:46	
Subject Sample	# 1	23:46	Subject Test Refused
Breath Volume	0.000	23:46	
Air Blank	0.000	23:47	
Control Test	0.080	23:47	
Air Blank	0.000	23:47	
Diagnostics Check	OK	23:47	

Data Download Date/Time: 06/26/2019 19:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,CHRISTOPHER,
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: MERHAR,MICHAEL
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 06/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:34

SUBJECT'S LAST NAME: WEDDING

DATE/TIME OF TEST: 06/28/2019 19:57:03

Diagnostics Check	OK	19:54	
Air Blank	0.000	19:54	
Control Test	0.079	19:55	
Air Blank	0.000	19:55	
Subject Sample	# 1	19:55	Subject Test Refused
Breath Volume	0.000	19:55	
Air Blank	0.000	19:56	
Control Test	0.079	19:56	
Air Blank	0.000	19:56	
Diagnostics Check	OK	19:57	

Data Download Date/Time: 07/21/2019 23:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

INSTRUMENT REGISTERED TO: NAS MAYPORT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOOS,JAMES,P
BREATH TEST OPERATOR AGENCY: Mayport Naval Base
ARRESTING OFFICER: LOPEZ,RAYMOND
ARRESTING OFFICER AGENCY: Mayport Naval Base
LAST AGENCY INSPECTION: 05/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: ALOMAR

DATE/TIME OF TEST: 06/12/2019 02:09:26

Diagnostics Check	OK	02:02
Air Blank	0.000	02:02
Control Test	0.078	02:03
Air Blank	0.000	02:03
Subject Sample # 1	0.122	02:04
Breath Volume	3.238	02:04
Air Blank	0.000	02:05
Air Blank	0.000	02:07
Subject Sample # 2	0.119	02:07
Breath Volume	2.355	02:07
Air Blank	0.000	02:08
Control Test	0.078	02:08
Air Blank	0.000	02:09
Diagnostics Check	OK	02:09

Data Download Date/Time: 06/13/2019 12:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

INSTRUMENT REGISTERED TO: NAS MAYPORT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHAHEEN,KEITH,B
BREATH TEST OPERATOR AGENCY: Mayport Naval Base
ARRESTING OFFICER: PARKER,DARBY
ARRESTING OFFICER AGENCY: Mayport Naval Base
LAST AGENCY INSPECTION: 05/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: ADMIN
ARREST TIME: 10:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:37

SUBJECT'S LAST NAME: HUDSON

DATE/TIME OF TEST: 06/12/2019 11:21:16

Diagnostics Check	OK	11:20	
Air Blank		11:20	RFI Detect
Air Blank	0.000	11:21	

Data Download Date/Time: 06/13/2019 12:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

INSTRUMENT REGISTERED TO: NAS MAYPORT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHAHEEN,KEITH,B
BREATH TEST OPERATOR AGENCY: Mayport Naval Base
ARRESTING OFFICER: PARKER,DARBY
ARRESTING OFFICER AGENCY: Mayport Naval Base
LAST AGENCY INSPECTION: 05/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: ADMIN
ARREST TIME: 10:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:37

SUBJECT'S LAST NAME: HUDSON

DATE/TIME OF TEST: 06/12/2019 11:56:12

Diagnostics Check	OK	11:50
Air Blank	0.000	11:50
Control Test	0.079	11:50
Air Blank	0.000	11:51
Subject Sample # 1	0.101	11:51
Breath Volume	4.964	11:51
Air Blank	0.000	11:52
Air Blank	0.000	11:54
Subject Sample # 2	0.097	11:54
Breath Volume	4.683	11:54
Air Blank	0.000	11:55
Control Test	0.078	11:55
Air Blank	0.000	11:56
Diagnostics Check	OK	11:56

Data Download Date/Time: 06/13/2019 12:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANER,OSVALDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KERN,ETHAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 06/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: VILLA-MORA

DATE/TIME OF TEST: 06/23/2019 01:33:22

Diagnostics Check		OK	01:27
Air Blank		0.000	01:27
Control Test		0.081	01:27
Air Blank		0.000	01:28
Subject Sample	# 1	0.262	01:29
Breath Volume		1.820	01:29
Air Blank		0.000	01:29
Air Blank		0.000	01:31
Subject Sample	# 2	0.262	01:31
Breath Volume		2.476	01:31
Air Blank		0.000	01:32
Control Test		0.080	01:32
Air Blank		0.000	01:33
Diagnostics Check		OK	01:33

Data Download Date/Time: 06/27/2019 13:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN, WILLIAM, L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: IRWIN, CHIP
ARRESTING OFFICER AGENCY: Windermere PD
LAST AGENCY INSPECTION: 06/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: OWEN

DATE/TIME OF TEST: 06/28/2019 01:31:31

Diagnostics Check	OK	01:24
Air Blank	0.000	01:24
Control Test	0.080	01:25
Air Blank	0.000	01:25
Subject Sample # 1	0.194	01:26
Breath Volume	3.515	01:26
Air Blank	0.000	01:27
Air Blank	0.000	01:29
Subject Sample # 2	0.201	01:29
Breath Volume	2.406	01:29
Air Blank	0.000	01:30
Control Test	0.080	01:30
Air Blank	0.000	01:31
Diagnostics Check	OK	01:31

Data Download Date/Time: 07/04/2019 16:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RIVERA,HERIBERTO
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 05/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: ROZANSKI

DATE/TIME OF TEST: 06/01/2019 00:05:35

Diagnostics Check	OK	23:59
Air Blank	0.000	00:00
Control Test	0.081	00:00
Air Blank	0.000	00:00
Subject Sample # 1	0.000	00:01
Breath Volume	4.320	00:01
Air Blank	0.000	00:01
Air Blank	0.000	00:03
Subject Sample # 2	0.000	00:04
Breath Volume	4.476	00:04
Air Blank	0.000	00:04
Control Test	0.081	00:05
Air Blank	0.000	00:05
Diagnostics Check	OK	00:05

Data Download Date/Time: 06/13/2019 14:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LANE,KEATON
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 05/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: LOPEZGOMEZ

DATE/TIME OF TEST: 06/01/2019 01:34:01

Diagnostics Check	OK	01:26
Air Blank	0.000	01:27
Control Test	0.081	01:27
Air Blank	0.000	01:28
Subject Sample # 1	0.156	01:29
Breath Volume	1.683	01:29
Air Blank	0.000	01:30
Air Blank	0.000	01:32
Subject Sample # 2	0.153	01:32
Breath Volume	1.605	01:32
Air Blank	0.000	01:33
Control Test	0.080	01:33
Air Blank	0.000	01:33
Diagnostics Check	OK	01:34

Data Download Date/Time: 06/13/2019 14:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MARTINEZ,ERNESTO
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: GARCIAFRANCO

DATE/TIME OF TEST: 06/01/2019 03:19:45

Diagnostics Check		OK	03:13
Air Blank		0.000	03:13
Control Test		0.081	03:14
Air Blank		0.000	03:14
Subject Sample	# 1	0.228	03:15
Breath Volume		2.632	03:15
Air Blank		0.000	03:16
Air Blank		0.000	03:17
Subject Sample	# 2	0.234	03:18
Breath Volume		2.105	03:18
Air Blank		0.000	03:18
Control Test		0.081	03:19
Air Blank		0.000	03:19
Diagnostics Check		OK	03:19

Data Download Date/Time: 06/13/2019 14:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MILLER,RYAN
ARRESTING OFFICER AGENCY: Windermere PD
LAST AGENCY INSPECTION: 05/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: GAUTHIER

DATE/TIME OF TEST: 06/01/2019 03:48:59

Diagnostics Check	OK	03:42
Air Blank	0.000	03:42
Control Test	0.081	03:43
Air Blank	0.000	03:43
Subject Sample # 1	0.081	03:43
Breath Volume	1.441	03:43
Air Blank	0.000	03:44
Air Blank	0.000	03:46
Subject Sample # 2	0.086	03:47
Breath Volume	1.484	03:47
Air Blank	0.000	03:48
Control Test	0.081	03:48
Air Blank	0.000	03:48
Diagnostics Check	OK	03:48

Data Download Date/Time: 06/13/2019 14:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SAUNDERS,HUNTER
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 05/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:51

SUBJECT'S LAST NAME: WARD

DATE/TIME OF TEST: 06/02/2019 00:22:13

Diagnostics Check		OK	00:15
Air Blank		0.000	00:15
Control Test		0.080	00:16
Air Blank		0.000	00:16
Subject Sample	# 1	0.212	00:17
Breath Volume		1.917	00:17
Air Blank		0.000	00:18
Air Blank		0.000	00:20
Subject Sample	# 2	0.218	00:20
Breath Volume		1.898	00:20
Air Blank		0.000	00:21
Control Test		0.080	00:21
Air Blank		0.000	00:22
Diagnostics Check		OK	00:22

Data Download Date/Time: 06/13/2019 14:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HARRISON,ROBBIE
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: PACHECO MONSALVE

DATE/TIME OF TEST: 06/02/2019 03:51:40

Diagnostics Check	OK	03:44
Air Blank	0.000	03:44
Control Test	0.081	03:44
Air Blank	0.000	03:45
Subject Sample # 1	0.173	03:47
Breath Volume	1.210	03:47
Air Blank	0.000	03:47
Air Blank	0.000	03:49
Subject Sample # 2	0.177	03:50
Breath Volume	1.144	03:50
Air Blank	0.000	03:50
Control Test	0.080	03:51
Air Blank	0.000	03:51
Diagnostics Check	OK	03:51

Data Download Date/Time: 06/13/2019 14:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RIDEAUX,ADAM
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: AUGUSTE

DATE/TIME OF TEST: 06/02/2019 05:05:11

Diagnostics Check	OK	04:55
Air Blank	0.000	04:55
Control Test	0.081	04:55
Air Blank	0.000	04:56
Subject Sample # 1	0.072	04:58
Breath Volume	1.449	04:58
Air Blank	0.000	04:58
Air Blank	0.000	05:00
Subject Sample # 2	0.069	05:03
Breath Volume	1.214	05:03
Air Blank	0.000	05:04
Control Test	0.081	05:04
Air Blank	0.000	05:05
Diagnostics Check	OK	05:05

Data Download Date/Time: 06/13/2019 14:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HOLLEY,DEANDRE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: GARCIAGONZALEZ

DATE/TIME OF TEST: 06/03/2019 00:48:08

Diagnostics Check	OK	00:45	
Air Blank	0.000	00:45	
Control Test	0.081	00:46	
Air Blank	0.000	00:46	
Subject Sample	# 1	00:46	Subject Test Refused
Breath Volume	0.000	00:46	
Air Blank	0.000	00:47	
Control Test	0.081	00:47	
Air Blank	0.000	00:48	
Diagnostics Check	OK	00:48	

Data Download Date/Time: 06/13/2019 14:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AUSTIN,ADAM
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 12:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:20

SUBJECT'S LAST NAME: COHEN

DATE/TIME OF TEST: 06/03/2019 12:49:57

Diagnostics Check	OK	12:43
Air Blank	0.000	12:44
Control Test	0.081	12:44
Air Blank	0.000	12:45
Subject Sample # 1	0.232	12:45
Breath Volume	1.882	12:45
Air Blank	0.000	12:46
Air Blank	0.000	12:47
Subject Sample # 2	0.228	12:48
Breath Volume	1.761	12:48
Air Blank	0.000	12:49
Control Test	0.080	12:49
Air Blank	0.000	12:49
Diagnostics Check	OK	12:49

Data Download Date/Time: 06/13/2019 14:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HARMAN, RICHARD
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: KERR

DATE/TIME OF TEST: 06/03/2019 23:54:41

Diagnostics Check	OK	23:46
Air Blank	0.000	23:46
Control Test	0.081	23:47
Air Blank	0.000	23:47
Subject Sample # 1	0.161	23:50
Breath Volume	1.250	23:50
Air Blank	0.000	23:50
Air Blank	0.000	23:52
Subject Sample # 2	0.157	23:53
Breath Volume	1.382	23:53
Air Blank	0.000	23:53
Control Test	0.080	23:54
Air Blank	0.000	23:54
Diagnostics Check	OK	23:54

Data Download Date/Time: 06/13/2019 14:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MARTELL, SIONE
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 05/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: FRESSE

DATE/TIME OF TEST: 06/05/2019 02:10:29

Diagnostics Check	OK	02:04
Air Blank	0.000	02:04
Control Test	0.081	02:05
Air Blank	0.000	02:05
Subject Sample # 1	0.091	02:06
Breath Volume	1.242	02:06
Air Blank	0.000	02:06
Air Blank	0.000	02:08
Subject Sample # 2	0.090	02:09
Breath Volume	3.167	02:09
Air Blank	0.000	02:09
Control Test	0.081	02:09
Air Blank	0.000	02:10
Diagnostics Check	OK	02:10

Data Download Date/Time: 06/13/2019 14:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HARRIS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 05/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: YOUNGS

DATE/TIME OF TEST: 06/07/2019 04:12:53

Diagnostics Check	OK	04:01	
Air Blank	0.000	04:01	
Control Test	0.081	04:02	
Air Blank	0.000	04:02	
Subject Sample	# 1	04:05	No Sample Provided
Breath Volume	0.000	04:05	
Air Blank	0.000	04:06	
Air Blank	0.000	04:08	
Subject Sample	# 2	04:11	No Sample Provided
Breath Volume	0.000	04:11	
Air Blank	0.000	04:11	
Control Test	0.081	04:12	
Air Blank	0.000	04:12	
Diagnostics Check	OK	04:12	

Data Download Date/Time: 06/13/2019 14:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANER,OSVALDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MILLER,RYAN
ARRESTING OFFICER AGENCY: Windermere PD
LAST AGENCY INSPECTION: 05/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: LIMA

DATE/TIME OF TEST: 06/07/2019 04:39:37

Diagnostics Check	OK	04:28	
Air Blank	0.000	04:28	
Control Test	0.080	04:29	
Air Blank	0.000	04:29	
Subject Sample	# 1	04:32	No Sample Provided
Breath Volume	0.000	04:32	
Air Blank	0.000	04:33	
Air Blank	0.000	04:35	
Subject Sample	# 2	04:38	No Sample Provided
Breath Volume	0.000	04:38	
Air Blank	0.000	04:38	
Control Test	0.081	04:39	
Air Blank	0.000	04:39	
Diagnostics Check	OK	04:39	

Data Download Date/Time: 06/13/2019 14:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HARRIS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 05/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: RAINER

DATE/TIME OF TEST: 06/07/2019 23:19:50

Diagnostics Check	OK	23:13
Air Blank	0.000	23:14
Control Test	0.081	23:14
Air Blank	0.000	23:14
Subject Sample # 1	0.124	23:15
Breath Volume	2.054	23:15
Air Blank	0.000	23:15
Air Blank	0.000	23:17
Subject Sample # 2	0.123	23:18
Breath Volume	1.714	23:18
Air Blank	0.000	23:18
Control Test	0.081	23:19
Air Blank	0.000	23:19
Diagnostics Check	OK	23:19

Data Download Date/Time: 06/13/2019 14:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: NDOUMBE,FREDERIC
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: WOODS

DATE/TIME OF TEST: 06/08/2019 03:51:37

Diagnostics Check		OK	03:35	
Air Blank		0.000	03:35	
Control Test		0.081	03:35	
Air Blank		0.000	03:36	
Subject Sample	# 1	0.132	03:38	
Breath Volume		1.507	03:38	
Air Blank		0.000	03:39	
Air Blank		0.000	03:41	
Subject Sample	# 2	0.113	03:44	Volume Not Met
Breath Volume		0.855	03:44	
Air Blank		0.000	03:45	
Air Blank		0.000	03:46	
Subject Sample	# 3	0.119	03:50	Volume Not Met
Breath Volume		0.390	03:50	
Air Blank		0.000	03:50	
Control Test		0.080	03:51	
Air Blank		0.000	03:51	
Diagnostics Check		OK	03:51	

Data Download Date/Time: 06/13/2019 14:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: NDOUMBE,FREDERIC
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: WOODS

DATE/TIME OF TEST: 06/08/2019 04:00:54

Diagnostics Check	OK	03:53
Air Blank	0.000	03:53
Control Test	0.079	03:54
Air Blank	0.000	03:54
Subject Sample # 1	0.117	03:55
Breath Volume	1.656	03:55
Air Blank	0.000	03:56
Air Blank	0.000	03:58
Subject Sample # 2	0.119	03:59
Breath Volume	1.605	03:59
Air Blank	0.000	04:00
Control Test	0.079	04:00
Air Blank	0.000	04:00
Diagnostics Check	OK	04:00

Data Download Date/Time: 06/13/2019 14:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LAWRENCE,NICKOLAS
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 05/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 06/08/2019 05:06:34

Diagnostics Check	OK	04:58
Air Blank	0.000	04:58
Control Test	0.081	04:59
Air Blank	0.000	04:59
Subject Sample # 1	0.174	05:02
Breath Volume	1.710	05:02
Air Blank	0.000	05:02
Air Blank	0.000	05:04
Subject Sample # 2	0.181	05:05
Breath Volume	1.500	05:05
Air Blank	0.000	05:05
Control Test	0.079	05:05
Air Blank	0.000	05:06
Diagnostics Check	OK	05:06

Data Download Date/Time: 06/13/2019 14:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KERN,ETHAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: ROSADO

DATE/TIME OF TEST: 06/09/2019 05:29:48

Diagnostics Check	OK	05:23
Air Blank	0.000	05:24
Control Test	0.081	05:24
Air Blank	0.000	05:24
Subject Sample # 1	0.158	05:25
Breath Volume	3.023	05:25
Air Blank	0.000	05:26
Air Blank	0.000	05:27
Subject Sample # 2	0.155	05:28
Breath Volume	3.121	05:28
Air Blank	0.000	05:28
Control Test	0.081	05:29
Air Blank	0.000	05:29
Diagnostics Check	OK	05:29

Data Download Date/Time: 06/13/2019 14:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TABBARA,RABIH
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 06:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:10

SUBJECT'S LAST NAME: ISHMAELJR

DATE/TIME OF TEST: 06/09/2019 07:45:06

Diagnostics Check		OK	07:37
Air Blank		0.000	07:38
Control Test		0.081	07:38
Air Blank		0.000	07:38
Subject Sample	# 1	0.224	07:39
Breath Volume		1.199	07:39
Air Blank		0.000	07:40
Air Blank		0.000	07:42
Subject Sample	# 2	0.220	07:43
Breath Volume		1.128	07:43
Air Blank		0.000	07:44
Control Test		0.081	07:44
Air Blank		0.000	07:45
Diagnostics Check		OK	07:45

Data Download Date/Time: 06/13/2019 14:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MICHAEL, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CISOWSKI, ANDREW
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 05/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 13:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:19

SUBJECT'S LAST NAME: MYRIE

DATE/TIME OF TEST: 06/09/2019 14:51:06

Diagnostics Check	OK	14:45
Air Blank	0.000	14:45
Control Test	0.078	14:45
Air Blank	0.000	14:46
Subject Sample # 1	0.298	14:46
Breath Volume	2.828	14:46
Air Blank	0.000	14:47
Air Blank	0.000	14:49
Subject Sample # 2	0.311	14:49
Breath Volume	3.062	14:49
Air Blank	0.000	14:50
Control Test	0.080	14:50
Air Blank	0.000	14:51
Diagnostics Check	OK	14:51

Data Download Date/Time: 06/13/2019 14:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MICHAEL, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: OCAMPO, OSCAR
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 05/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:40

SUBJECT'S LAST NAME: SANTOSHHERNANDEZ

DATE/TIME OF TEST: 06/09/2019 17:12:41

Diagnostics Check		OK	17:06
Air Blank		0.000	17:06
Control Test		0.080	17:07
Air Blank		0.000	17:07
Subject Sample	# 1	0.204	17:08
Breath Volume		2.457	17:08
Air Blank		0.000	17:09
Air Blank		0.000	17:10
Subject Sample	# 2	0.202	17:11
Breath Volume		3.105	17:11
Air Blank		0.000	17:11
Control Test		0.081	17:12
Air Blank		0.000	17:12
Diagnostics Check		OK	17:12

Data Download Date/Time: 06/13/2019 14:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RIVERA, HERIBERTO
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 05/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 06/12/2019 01:06:46

Diagnostics Check	OK	01:00
Air Blank	0.000	01:01
Control Test	0.080	01:01
Air Blank	0.000	01:02
Subject Sample # 1	0.012	01:02
Breath Volume	2.941	01:02
Air Blank	0.000	01:02
Air Blank	0.000	01:04
Subject Sample # 2	0.011	01:05
Breath Volume	2.019	01:05
Air Blank	0.000	01:05
Control Test	0.081	01:06
Air Blank	0.000	01:06
Diagnostics Check	OK	01:06

Data Download Date/Time: 06/13/2019 14:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANER,OSVALDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KISSANE,BRIAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: ACOSTA

DATE/TIME OF TEST: 06/12/2019 03:12:15

Diagnostics Check		OK	03:06
Air Blank		0.000	03:06
Control Test		0.081	03:06
Air Blank		0.000	03:07
Subject Sample	# 1	0.222	03:07
Breath Volume		1.761	03:07
Air Blank		0.000	03:08
Air Blank		0.000	03:10
Subject Sample	# 2	0.212	03:10
Breath Volume		1.406	03:10
Air Blank		0.000	03:11
Control Test		0.080	03:11
Air Blank		0.000	03:12
Diagnostics Check		OK	03:12

Data Download Date/Time: 06/13/2019 14:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MARTELL,SIONE
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 05/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: SKOOG

DATE/TIME OF TEST: 06/12/2019 04:17:47

Diagnostics Check		OK	04:11
Air Blank		0.000	04:12
Control Test		0.081	04:12
Air Blank		0.000	04:12
Subject Sample	# 1	0.109	04:13
Breath Volume		2.800	04:13
Air Blank		0.000	04:13
Air Blank		0.000	04:15
Subject Sample	# 2	0.107	04:16
Breath Volume		3.203	04:16
Air Blank		0.000	04:16
Control Test		0.080	04:17
Air Blank		0.000	04:17
Diagnostics Check		OK	04:17

Data Download Date/Time: 06/13/2019 14:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PRICE,FORREST
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: RIOSLAVERDE

DATE/TIME OF TEST: 06/13/2019 00:09:38

Diagnostics Check		OK	00:03
Air Blank		0.000	00:03
Control Test		0.081	00:03
Air Blank		0.000	00:04
Subject Sample	# 1	0.194	00:05
Breath Volume		1.738	00:05
Air Blank		0.000	00:05
Air Blank		0.000	00:07
Subject Sample	# 2	0.203	00:08
Breath Volume		2.636	00:08
Air Blank		0.000	00:08
Control Test		0.080	00:09
Air Blank		0.000	00:09
Diagnostics Check		OK	00:09

Data Download Date/Time: 06/13/2019 14:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KERN,ETHAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 06/13/2019 02:29:23

Diagnostics Check	OK	02:22
Air Blank	0.000	02:23
Control Test	0.081	02:23
Air Blank	0.000	02:23
Subject Sample # 1	0.000	02:24
Breath Volume	2.628	02:24
Air Blank	0.000	02:25
Air Blank	0.000	02:27
Subject Sample # 2	0.000	02:28
Breath Volume	2.730	02:28
Air Blank	0.000	02:28
Control Test	0.081	02:28
Air Blank	0.000	02:29
Diagnostics Check	OK	02:29

Data Download Date/Time: 06/13/2019 14:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANER, OSVALDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WALLACE, TIMOTHY
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 05/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: LUMBI-RODRIGUEZ

DATE/TIME OF TEST: 06/13/2019 21:46:50

Diagnostics Check	OK	21:40
Air Blank	0.000	21:41
Control Test	0.080	21:41
Air Blank	0.000	21:42
Subject Sample # 1	0.140	21:42
Breath Volume	1.406	21:42
Air Blank	0.000	21:43
Air Blank	0.000	21:45
Subject Sample # 2	0.138	21:45
Breath Volume	1.199	21:45
Air Blank	0.000	21:45
Control Test	0.080	21:46
Air Blank	0.000	21:46
Diagnostics Check	OK	21:46

Data Download Date/Time: 06/19/2019 10:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MARTELL,SIONE
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 05/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: GOMEZ RESTREPO

DATE/TIME OF TEST: 06/13/2019 23:05:33

Diagnostics Check	OK	22:59
Air Blank	0.000	23:00
Control Test	0.081	23:00
Air Blank	0.000	23:00
Subject Sample # 1	0.000	23:01
Breath Volume	5.105	23:01
Air Blank	0.000	23:01
Air Blank	0.000	23:03
Subject Sample # 2	0.000	23:04
Breath Volume	5.273	23:04
Air Blank	0.000	23:04
Control Test	0.081	23:04
Air Blank	0.000	23:05
Diagnostics Check	OK	23:05

Data Download Date/Time: 06/19/2019 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KERN,ETHAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 06/14/2019 03:06:48

Diagnostics Check	OK	02:59
Air Blank	0.000	02:59
Control Test	0.082	03:00
Air Blank	0.000	03:00
Subject Sample # 1	0.160	03:01
Breath Volume	1.878	03:01
Air Blank	0.000	03:02
Air Blank	0.000	03:04
Subject Sample # 2	0.155	03:05
Breath Volume	2.019	03:05
Air Blank	0.000	03:05
Control Test	0.081	03:06
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

Data Download Date/Time: 06/19/2019 10:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAM,ALEXANDER
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: DE JESUS GOMEZ

DATE/TIME OF TEST: 06/16/2019 01:18:12

Diagnostics Check	OK	01:11
Air Blank	0.000	01:11
Control Test	0.081	01:11
Air Blank	0.000	01:12
Subject Sample # 1	0.211	01:13
Breath Volume	2.125	01:13
Air Blank	0.000	01:14
Air Blank	0.000	01:16
Subject Sample # 2	0.206	01:16
Breath Volume	1.515	01:16
Air Blank	0.000	01:17
Control Test	0.081	01:17
Air Blank	0.000	01:18
Diagnostics Check	OK	01:18

Data Download Date/Time: 06/19/2019 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAWKINS,DEBORAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:08

SUBJECT'S LAST NAME: BOLANOS

DATE/TIME OF TEST: 06/16/2019 05:43:23

Diagnostics Check		OK	05:34
Air Blank		0.000	05:34
Control Test		0.081	05:34
Air Blank		0.000	05:35
Subject Sample	# 1	0.203	05:38
Breath Volume		1.199	05:38
Air Blank		0.000	05:38
Air Blank		0.000	05:40
Subject Sample	# 2	0.208	05:41
Breath Volume		1.144	05:41
Air Blank		0.000	05:42
Control Test		0.079	05:42
Air Blank		0.000	05:43
Diagnostics Check		OK	05:43

Data Download Date/Time: 06/19/2019 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HARMAN,RICHARD
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:27

SUBJECT'S LAST NAME: GUZMANCRUZ

DATE/TIME OF TEST: 06/16/2019 05:58:57

Diagnostics Check		OK	05:51
Air Blank		0.000	05:51
Control Test		0.080	05:51
Air Blank		0.000	05:52
Subject Sample	# 1	0.118	05:53
Breath Volume		1.363	05:53
Air Blank		0.000	05:54
Air Blank		0.000	05:55
Subject Sample	# 2	0.113	05:57
Breath Volume		1.316	05:57
Air Blank		0.000	05:58
Control Test		0.080	05:58
Air Blank		0.000	05:58
Diagnostics Check		OK	05:58

Data Download Date/Time: 06/19/2019 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GORDON,CURTIS
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:48

SUBJECT'S LAST NAME: COLONMARTINEZ

DATE/TIME OF TEST: 06/16/2019 06:23:23

Diagnostics Check	OK	06:16
Air Blank	0.000	06:16
Control Test	0.080	06:17
Air Blank	0.000	06:17
Subject Sample # 1	0.141	06:18
Breath Volume	2.203	06:18
Air Blank	0.000	06:19
Air Blank	0.000	06:21
Subject Sample # 2	0.138	06:21
Breath Volume	1.390	06:21
Air Blank	0.000	06:22
Control Test	0.080	06:22
Air Blank	0.000	06:23
Diagnostics Check	OK	06:23

Data Download Date/Time: 06/19/2019 10:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MILLER,JACOB
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 05/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:35

SUBJECT'S LAST NAME: RAMIREZGARCIA

DATE/TIME OF TEST: 06/16/2019 18:16:24

Diagnostics Check	OK	18:08
Air Blank	0.000	18:09
Control Test	0.081	18:09
Air Blank	0.000	18:10
Subject Sample # 1	0.215	18:11
Breath Volume	1.421	18:11
Air Blank	0.000	18:12
Air Blank	0.000	18:14
Subject Sample # 2	0.220	18:14
Breath Volume	1.578	18:14
Air Blank	0.000	18:15
Control Test	0.080	18:15
Air Blank	0.000	18:16
Diagnostics Check	OK	18:16

Data Download Date/Time: 06/19/2019 10:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: NGUYEN,TRANG
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:21

SUBJECT'S LAST NAME: HERRING

DATE/TIME OF TEST: 06/16/2019 19:22:09

Diagnostics Check		OK	19:07	
Air Blank		0.000	19:07	
Control Test		0.080	19:08	
Air Blank		0.000	19:08	
Subject Sample	# 1	0.214	19:11	Volume Not Met
Breath Volume		0.832	19:11	
Air Blank		0.000	19:12	
Air Blank		0.000	19:14	
Subject Sample	# 2	0.224	19:17	
Breath Volume		1.269	19:17	
Air Blank		0.000	19:17	
Air Blank		0.000	19:19	
Subject Sample	# 3	0.224	19:20	
Breath Volume		1.281	19:20	
Air Blank		0.000	19:21	
Control Test		0.079	19:21	
Air Blank		0.000	19:22	
Diagnostics Check		OK	19:22	

Data Download Date/Time: 06/19/2019 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AZEVEDO,FABIO
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: POLLARD

DATE/TIME OF TEST: 06/17/2019 01:26:22

Diagnostics Check	OK	01:20
Air Blank	0.000	01:20
Control Test	0.081	01:20
Air Blank	0.000	01:21
Subject Sample # 1	0.085	01:21
Breath Volume	2.339	01:21
Air Blank	0.000	01:22
Air Blank	0.000	01:24
Subject Sample # 2	0.083	01:24
Breath Volume	1.105	01:24
Air Blank	0.000	01:25
Control Test	0.081	01:25
Air Blank	0.000	01:26
Diagnostics Check	OK	01:26

Data Download Date/Time: 06/19/2019 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MICHAEL, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RODRIGUEZ, YESSICA
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 07:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:20

SUBJECT'S LAST NAME: ADDERLINIGARCIA

DATE/TIME OF TEST: 06/17/2019 08:50:39

Diagnostics Check	OK	08:44
Air Blank	0.000	08:45
Control Test	0.081	08:45
Air Blank	0.000	08:45
Subject Sample # 1	0.034	08:46
Breath Volume	3.152	08:46
Air Blank	0.000	08:46
Air Blank	0.000	08:48
Subject Sample # 2	0.032	08:49
Breath Volume	1.976	08:49
Air Blank	0.000	08:49
Control Test	0.080	08:50
Air Blank	0.000	08:50
Diagnostics Check	OK	08:50

Data Download Date/Time: 06/19/2019 10:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: NAS,JOSEPH
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: FRONK

DATE/TIME OF TEST: 06/18/2019 01:27:45

Diagnostics Check		OK	01:20
Air Blank		0.000	01:21
Control Test		0.081	01:21
Air Blank		0.000	01:22
Subject Sample	# 1	0.239	01:23
Breath Volume		2.312	01:23
Air Blank		0.000	01:23
Air Blank		0.000	01:25
Subject Sample	# 2	0.247	01:26
Breath Volume		2.843	01:26
Air Blank		0.000	01:26
Control Test		0.080	01:27
Air Blank		0.000	01:27
Diagnostics Check		OK	01:27

Data Download Date/Time: 06/19/2019 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VENTURA, CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:54

SUBJECT'S LAST NAME: CAREY-ROSS

DATE/TIME OF TEST: 06/18/2019 05:22:30

Diagnostics Check	OK	05:16
Air Blank	0.000	05:16
Control Test	0.081	05:17
Air Blank	0.000	05:17
Subject Sample # 1	0.000	05:18
Breath Volume	2.156	05:18
Air Blank	0.000	05:18
Air Blank	0.000	05:20
Subject Sample # 2	0.000	05:21
Breath Volume	2.093	05:21
Air Blank	0.000	05:21
Control Test	0.081	05:21
Air Blank	0.000	05:22
Diagnostics Check	OK	05:22

Data Download Date/Time: 06/19/2019 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HARRIS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 05/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: GAONA PEREZ

DATE/TIME OF TEST: 06/18/2019 23:26:36

Diagnostics Check	OK	23:20
Air Blank	0.000	23:20
Control Test	0.081	23:21
Air Blank	0.000	23:21
Subject Sample # 1	0.094	23:22
Breath Volume	4.171	23:22
Air Blank	0.000	23:22
Air Blank	0.000	23:24
Subject Sample # 2	0.095	23:25
Breath Volume	4.089	23:25
Air Blank	0.000	23:25
Control Test	0.081	23:26
Air Blank	0.000	23:26
Diagnostics Check	OK	23:26

Data Download Date/Time: 06/19/2019 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: FLOOD,GERARD
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 05/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: SORRELL

DATE/TIME OF TEST: 06/18/2019 23:45:41

Diagnostics Check	OK	23:39
Air Blank	0.000	23:40
Control Test	0.080	23:40
Air Blank	0.000	23:40
Subject Sample # 1	0.217	23:41
Breath Volume	1.664	23:41
Air Blank	0.000	23:41
Air Blank	0.000	23:43
Subject Sample # 2	0.208	23:44
Breath Volume	1.671	23:44
Air Blank	0.000	23:44
Control Test	0.080	23:45
Air Blank	0.000	23:45
Diagnostics Check	OK	23:45

Data Download Date/Time: 06/19/2019 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HARMAN, RICHARD
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: NICHOLAS

DATE/TIME OF TEST: 06/19/2019 02:43:36

Diagnostics Check		OK	02:36
Air Blank		0.000	02:37
Control Test		0.081	02:37
Air Blank		0.000	02:37
Subject Sample	# 1	0.288	02:38
Breath Volume		1.328	02:38
Air Blank		0.000	02:39
Air Blank		0.000	02:40
Subject Sample	# 2	0.283	02:42
Breath Volume		2.937	02:42
Air Blank		0.000	02:42
Control Test		0.080	02:42
Air Blank		0.000	02:43
Diagnostics Check		OK	02:43

Data Download Date/Time: 06/19/2019 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BONK,JAYSON
ARRESTING OFFICER AGENCY: Windermere PD
LAST AGENCY INSPECTION: 05/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 06/19/2019 03:58:19

Diagnostics Check	OK	03:52
Air Blank	0.000	03:52
Control Test	0.079	03:53
Air Blank	0.000	03:53
Subject Sample # 1	0.264	03:54
Breath Volume	1.585	03:54
Air Blank	0.000	03:54
Air Blank	0.000	03:56
Subject Sample # 2	0.260	03:56
Breath Volume	2.757	03:56
Air Blank	0.000	03:57
Control Test	0.080	03:57
Air Blank	0.000	03:58
Diagnostics Check	OK	03:58

Data Download Date/Time: 06/19/2019 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BINET, MARLON
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: HEBERT

DATE/TIME OF TEST: 06/19/2019 06:31:14

Diagnostics Check	OK	06:24
Air Blank	0.000	06:24
Control Test	0.080	06:25
Air Blank	0.000	06:25
Subject Sample # 1	0.135	06:26
Breath Volume	2.347	06:26
Air Blank	0.000	06:27
Air Blank	0.000	06:29
Subject Sample # 2	0.134	06:29
Breath Volume	1.363	06:29
Air Blank	0.000	06:30
Control Test	0.080	06:30
Air Blank	0.000	06:31
Diagnostics Check	OK	06:31

Data Download Date/Time: 06/19/2019 10:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001442

INSTRUMENT REGISTERED TO: BISCAYNE NAT PARK
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HELMERS,DANIEL,W
BREATH TEST OPERATOR AGENCY: BISCAYNE NATL PARK
ARRESTING OFFICER: RUIZ,RUIZ
ARRESTING OFFICER AGENCY: BISCAYNE NATL PARK
LAST AGENCY INSPECTION: 05/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: BUI
ARREST TIME: 15:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: IGLESIAS

DATE/TIME OF TEST: 06/23/2019 16:36:21

Diagnostics Check	OK	16:33	
Air Blank	0.000	16:33	
Control Test	0.078	16:34	
Air Blank	0.000	16:34	
Subject Sample	# 1	16:35	Subject Test Refused
Breath Volume	0.000	16:35	
Air Blank	0.000	16:35	
Control Test	0.079	16:35	
Air Blank	0.000	16:36	
Diagnostics Check	OK	16:36	

Data Download Date/Time: 06/27/2019 09:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTSTATT,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: CRICK,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 05/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: NGUYEN

DATE/TIME OF TEST: 06/09/2019 04:38:48

Diagnostics Check		OK	04:27	
Air Blank		0.000	04:27	
Control Test		0.081	04:28	
Air Blank		0.000	04:28	
Subject Sample	# 1	0.000	04:31	Volume Not Met
Breath Volume		0.019	04:31	
Air Blank		0.000	04:32	
Air Blank		0.000	04:34	
Subject Sample	# 2	0.062	04:37	Volume Not Met
Breath Volume		0.257	04:37	
Air Blank		0.000	04:37	
Control Test		0.082	04:38	
Air Blank		0.000	04:38	
Diagnostics Check		OK	04:38	

Data Download Date/Time: 06/25/2019 09:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTSTATT, MATTHEW, B
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: CRICK, M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 05/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: NGUYEN

DATE/TIME OF TEST: 06/09/2019 05:23:09

Diagnostics Check		OK	05:09	
Air Blank		0.000	05:10	
Control Test		0.081	05:10	
Air Blank		0.000	05:10	
Subject Sample	# 1	0.063	05:14	Volume Not Met
Breath Volume		0.859	05:14	
Air Blank		0.000	05:14	
Air Blank		0.000	05:16	
Subject Sample	# 2	0.072	05:18	
Breath Volume		2.109	05:18	
Air Blank		0.000	05:19	
Air Blank		0.000	05:21	
Subject Sample	# 3	0.061	05:21	
Breath Volume		1.449	05:21	
Air Blank		0.000	05:22	
Control Test		0.080	05:22	
Air Blank		0.000	05:23	
Diagnostics Check		OK	05:23	

Data Download Date/Time: 06/25/2019 09:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTORO,CAMERON,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 05/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: BEAVERS

DATE/TIME OF TEST: 06/09/2019 19:13:20

Diagnostics Check	OK	19:10	
Air Blank	0.000	19:10	
Control Test	0.081	19:11	
Air Blank	0.000	19:11	
Subject Sample	# 1	19:11	Subject Test Refused
Breath Volume	0.000	19:11	
Air Blank	0.000	19:12	
Control Test	0.081	19:12	
Air Blank	0.000	19:13	
Diagnostics Check	OK	19:13	

Data Download Date/Time: 06/25/2019 09:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENTZ,JARROD,C
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: LOVETT,K
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 05/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: FLICKINGER

DATE/TIME OF TEST: 06/10/2019 20:46:04

Diagnostics Check	OK	20:44	
Air Blank	0.000	20:45	
Control Test	0.000	20:45	RFI Detect
Air Blank	0.000	20:46	

Data Download Date/Time: 06/25/2019 09:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENTZ,JARROD,C
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: LOVETT,K
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 05/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: FLICKINGER

DATE/TIME OF TEST: 06/10/2019 20:51:48

Diagnostics Check	OK	20:49	
Air Blank	0.000	20:49	
Control Test	0.081	20:49	
Air Blank	0.000	20:50	
Subject Sample	# 1	20:50	Subject Test Refused
Breath Volume	0.000	20:50	
Air Blank	0.000	20:50	
Control Test	0.081	20:51	
Air Blank	0.000	20:51	
Diagnostics Check	OK	20:51	

Data Download Date/Time: 06/25/2019 09:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARLOW,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: PRITCHARD,TRAVIS
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 05/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 06/11/2019 18:57:41

Diagnostics Check	OK	18:50
Air Blank	0.000	18:51
Control Test	0.081	18:51
Air Blank	0.000	18:51
Subject Sample # 1	0.000	18:52
Breath Volume	1.375	18:52
Air Blank	0.000	18:53
Air Blank	0.000	18:55
Subject Sample # 2	0.000	18:56
Breath Volume	1.375	18:56
Air Blank	0.000	18:56
Control Test	0.080	18:57
Air Blank	0.000	18:57
Diagnostics Check	OK	18:57

Data Download Date/Time: 06/25/2019 09:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARCHER,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 05/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 06/13/2019 03:51:53

Diagnostics Check	OK	03:43	
Air Blank	0.000	03:43	
Control Test	0.080	03:44	
Air Blank	0.000	03:44	
Subject Sample	# 1	03:47	No Sample Provided
Breath Volume	0.000	03:47	
Air Blank	0.000	03:48	
Air Blank	0.000	03:50	
Subject Sample	# 2	03:50	Subject Test Refused
Breath Volume	0.000	03:50	
Air Blank	0.000	03:50	
Control Test	0.081	03:51	
Air Blank	0.000	03:51	
Diagnostics Check	OK	03:51	

Data Download Date/Time: 06/25/2019 09:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASLINE, T'NEISHA, S
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY, R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 05/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: LANHAM

DATE/TIME OF TEST: 06/14/2019 04:59:26

Diagnostics Check		OK	04:53
Air Blank		0.000	04:53
Control Test		0.081	04:53
Air Blank		0.000	04:54
Subject Sample	# 1	0.263	04:54
Breath Volume		1.511	04:54
Air Blank		0.000	04:55
Air Blank		0.000	04:57
Subject Sample	# 2	0.243	04:57
Breath Volume		1.410	04:57
Air Blank		0.000	04:58
Control Test		0.080	04:58
Air Blank		0.000	04:59
Diagnostics Check		OK	04:59

Data Download Date/Time: 06/25/2019 09:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENTZ,JARROD,C
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: LOVETT,K
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 05/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:19

SUBJECT'S LAST NAME: CARTER-BAILEY

DATE/TIME OF TEST: 06/14/2019 20:54:50

Diagnostics Check	OK	20:43
Air Blank	0.000	20:43
Control Test	0.081	20:43
Air Blank	0.000	20:44
Subject Sample # 1	0.279	20:44
Breath Volume	2.089	20:44
Air Blank	0.000	20:45
Air Blank	0.000	20:47
Subject Sample # 2	0.252	20:48
Breath Volume	1.183	20:48
Air Blank	0.000	20:49
Air Blank	0.000	20:50
Subject Sample # 3	0.270	20:53
Breath Volume	1.976	20:53
Air Blank	0.000	20:53
Control Test	0.081	20:54
Air Blank	0.000	20:54
Diagnostics Check	OK	20:54

Data Download Date/Time: 06/25/2019 09:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARLOW,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: TAPAWAN,MICHAEL
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 05/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 06/15/2019 01:11:22

Diagnostics Check	OK	01:02	
Air Blank	0.000	01:02	
Control Test	0.080	01:03	
Air Blank	0.000	01:03	
Subject Sample	# 1	01:06	No Sample Provided
Breath Volume	0.000	01:06	
Air Blank	0.000	01:07	
Air Blank	0.000	01:09	
Subject Sample	# 2	01:10	Subject Test Refused
Breath Volume	0.000	01:10	
Air Blank	0.000	01:10	
Control Test	0.081	01:10	
Air Blank	0.000	01:11	
Diagnostics Check	OK	01:11	

Data Download Date/Time: 06/25/2019 09:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENTZ,JARROD,C
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: PRITCHARD,T
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 05/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: SAAVEDRA

DATE/TIME OF TEST: 06/15/2019 04:12:12

Diagnostics Check		OK	03:59	
Air Blank		0.000	03:59	
Control Test		0.081	04:00	
Air Blank		0.000	04:00	
Subject Sample	# 1	0.111	04:03	Volume Not Met
Breath Volume		0.574	04:03	
Air Blank		0.000	04:04	
Air Blank		0.000	04:06	
Subject Sample	# 2	0.135	04:06	
Breath Volume		1.402	04:06	
Air Blank		0.000	04:07	
Air Blank		0.000	04:09	
Subject Sample	# 3	0.136	04:10	
Breath Volume		1.300	04:10	
Air Blank		0.000	04:11	
Control Test		0.080	04:11	
Air Blank		0.000	04:12	
Diagnostics Check		OK	04:12	

Data Download Date/Time: 06/25/2019 09:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARLOW,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: LOVETT,KEITH
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 05/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 06/15/2019 04:41:21

Diagnostics Check	OK	04:32	
Air Blank	0.000	04:33	
Control Test	0.080	04:33	
Air Blank	0.000	04:33	
Subject Sample	# 1	04:36	No Sample Provided
Breath Volume	0.000	04:36	
Air Blank	0.000	04:37	
Air Blank	0.000	04:39	
Subject Sample	# 2	04:40	Subject Test Refused
Breath Volume	0.000	04:40	
Air Blank	0.000	04:40	
Control Test	0.081	04:40	
Air Blank	0.000	04:41	
Diagnostics Check	OK	04:41	

Data Download Date/Time: 06/25/2019 09:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTSTATT,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: HOULE,A
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 05/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 06/18/2019 00:09:21

Diagnostics Check	OK	00:00	
Air Blank	0.000	00:01	
Control Test	0.081	00:01	
Air Blank	0.000	00:02	
Subject Sample	# 1	00:05	No Sample Provided
Breath Volume	0.000	00:05	
Air Blank	0.000	00:05	
Air Blank	0.000	00:07	
Subject Sample	# 2	00:08	Subject Test Refused
Breath Volume	0.000	00:08	
Air Blank	0.000	00:08	
Control Test	0.081	00:08	
Air Blank	0.000	00:09	
Diagnostics Check	OK	00:09	

Data Download Date/Time: 06/25/2019 09:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARCHER,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: LOVETT,K
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 05/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: WOODLEY

DATE/TIME OF TEST: 06/19/2019 03:49:35

Diagnostics Check	OK	03:43
Air Blank	0.000	03:43
Control Test	0.081	03:43
Air Blank	0.000	03:44
Subject Sample # 1	0.152	03:45
Breath Volume	3.800	03:45
Air Blank	0.000	03:45
Air Blank	0.000	03:47
Subject Sample # 2	0.151	03:48
Breath Volume	3.902	03:48
Air Blank	0.000	03:48
Control Test	0.081	03:49
Air Blank	0.000	03:49
Diagnostics Check	OK	03:49

Data Download Date/Time: 06/25/2019 09:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARLOW,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: SEWELL,ANDRE
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 05/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 06/24/2019 21:35:58

Diagnostics Check	OK	21:29
Air Blank	0.000	21:30
Control Test	0.080	21:30
Air Blank	0.000	21:31
Subject Sample # 1	0.000	21:31
Breath Volume	1.867	21:31
Air Blank	0.000	21:32
Air Blank	0.000	21:34
Subject Sample # 2	0.000	21:34
Breath Volume	1.601	21:34
Air Blank	0.000	21:35
Control Test	0.079	21:35
Air Blank	0.000	21:35
Diagnostics Check	OK	21:35

Data Download Date/Time: 06/25/2019 09:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARCHER,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:38

SUBJECT'S LAST NAME: STRICKLAND

DATE/TIME OF TEST: 06/26/2019 18:13:44

Diagnostics Check	OK	18:05
Air Blank	0.000	18:05
Control Test	0.081	18:06
Air Blank	0.000	18:06
Subject Sample # 1	0.000	18:08
Breath Volume	2.042	18:08
Air Blank	0.000	18:08
Air Blank	0.000	18:10
Subject Sample # 2	0.000	18:12
Breath Volume	1.859	18:12
Air Blank	0.000	18:12
Control Test	0.080	18:13
Air Blank	0.000	18:13
Diagnostics Check	OK	18:13

Data Download Date/Time: 07/25/2019 11:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENTZ,JARROD,C
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: ADAMS,S
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: GORDON

DATE/TIME OF TEST: 06/28/2019 19:54:21

Diagnostics Check	OK	19:48
Air Blank	0.000	19:48
Control Test	0.081	19:48
Air Blank	0.000	19:49
Subject Sample # 1	0.252	19:49
Breath Volume	1.824	19:49
Air Blank	0.000	19:50
Air Blank	0.000	19:52
Subject Sample # 2	0.245	19:52
Breath Volume	2.054	19:52
Air Blank	0.000	19:53
Control Test	0.080	19:53
Air Blank	0.000	19:54
Diagnostics Check	OK	19:54

Data Download Date/Time: 07/25/2019 11:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARLOW,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: PRITCHARD,TRAVIS
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:59

SUBJECT'S LAST NAME: FAIRCLOTH

DATE/TIME OF TEST: 06/29/2019 21:38:54

Diagnostics Check	OK	21:27	
Air Blank	0.000	21:28	
Control Test	0.081	21:28	
Air Blank	0.000	21:28	
Subject Sample	# 1	21:31	No Sample Provided
Breath Volume	0.000	21:31	
Air Blank	0.000	21:32	
Air Blank	0.000	21:34	
Subject Sample	# 2	21:37	No Sample Provided
Breath Volume	0.000	21:37	
Air Blank	0.000	21:38	
Control Test	0.081	21:38	
Air Blank	0.000	21:38	
Diagnostics Check	OK	21:38	

Data Download Date/Time: 07/25/2019 11:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARLOW,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: LINEBERRY,SCOTT
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 06/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: CRUMPLER

DATE/TIME OF TEST: 06/30/2019 04:14:29

Diagnostics Check	OK	04:05	
Air Blank	0.000	04:05	
Control Test	0.081	04:06	
Air Blank	0.000	04:06	
Subject Sample	# 1	04:09	No Sample Provided
Breath Volume	0.000	04:09	
Air Blank	0.000	04:10	
Air Blank	0.000	04:12	
Subject Sample	# 2	04:13	Subject Test Refused
Breath Volume	0.000	04:13	
Air Blank	0.000	04:13	
Control Test	0.081	04:13	
Air Blank	0.000	04:14	
Diagnostics Check	OK	04:14	

Data Download Date/Time: 07/25/2019 11:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINI,RYAN,H
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: ENGH,WILLIAM
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: MUCKENSTURM

DATE/TIME OF TEST: 06/01/2019 23:00:13

Diagnostics Check	OK	22:52
Air Blank	0.000	22:53
Control Test	0.081	22:53
Air Blank	0.000	22:53
Subject Sample # 1	0.191	22:55
Breath Volume	1.796	22:55
Air Blank	0.000	22:55
Air Blank	0.000	22:57
Subject Sample # 2	0.188	22:58
Breath Volume	1.484	22:58
Air Blank	0.000	22:59
Control Test	0.079	22:59
Air Blank	0.000	23:00
Diagnostics Check	OK	23:00

Data Download Date/Time: 06/26/2019 12:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINI,RYAN,H
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: BROWN,SAMANTHA
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 05:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:37

SUBJECT'S LAST NAME: DEEMS

DATE/TIME OF TEST: 06/06/2019 06:04:47

Diagnostics Check	OK	05:58
Air Blank	0.000	05:59
Control Test	0.081	05:59
Air Blank	0.000	06:00
Subject Sample # 1	0.000	06:00
Breath Volume	3.195	06:00
Air Blank	0.000	06:00
Air Blank	0.000	06:03
Subject Sample # 2	0.000	06:03
Breath Volume	3.000	06:03
Air Blank	0.000	06:03
Control Test	0.081	06:04
Air Blank	0.000	06:04
Diagnostics Check	OK	06:04

Data Download Date/Time: 06/26/2019 12:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZARRILLO,N,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: LOWE,A
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 06/09/2019 03:01:55

Diagnostics Check	OK	02:58	
Air Blank	0.000	02:58	
Control Test	0.081	02:58	
Air Blank	0.000	02:59	
Subject Sample	# 1	03:00	Subject Test Refused
Breath Volume	0.000	03:00	
Air Blank	0.000	03:01	
Control Test	0.081	03:01	
Air Blank	0.000	03:01	
Diagnostics Check	OK	03:01	

Data Download Date/Time: 06/26/2019 12:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: ZARRILLO,NICHOLAS
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: PELLEY

DATE/TIME OF TEST: 06/10/2019 03:45:13

Diagnostics Check	OK	03:36
Air Blank	0.000	03:36
Control Test	0.080	03:37
Air Blank	0.000	03:37
Subject Sample # 1	0.159	03:40
Breath Volume	3.242	03:40
Air Blank	0.000	03:40
Air Blank	0.000	03:42
Subject Sample # 2	0.156	03:43
Breath Volume	1.660	03:43
Air Blank	0.000	03:44
Control Test	0.078	03:44
Air Blank	0.000	03:45
Diagnostics Check	OK	03:45

Data Download Date/Time: 06/26/2019 12:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: GRENNAN

DATE/TIME OF TEST: 06/13/2019 03:25:57

Diagnostics Check	OK	03:19
Air Blank	0.000	03:20
Control Test	0.080	03:20
Air Blank	0.000	03:21
Subject Sample # 1	0.102	03:21
Breath Volume	3.117	03:21
Air Blank	0.000	03:22
Air Blank	0.000	03:23
Subject Sample # 2	0.100	03:24
Breath Volume	1.675	03:24
Air Blank	0.000	03:25
Control Test	0.080	03:25
Air Blank	0.000	03:25
Diagnostics Check	OK	03:25

Data Download Date/Time: 06/26/2019 12:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FUGIEL,ADRIAN,M
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: SANCHO,BRANDON
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: MANGES

DATE/TIME OF TEST: 06/14/2019 23:00:56

Diagnostics Check	OK	22:58	
Air Blank	0.000	22:58	
Control Test	0.080	22:58	
Air Blank	0.000	22:59	
Subject Sample	# 1	22:59	Subject Test Refused
Breath Volume	0.000	22:59	
Air Blank	0.000	23:00	
Control Test	0.079	23:00	
Air Blank	0.000	23:00	
Diagnostics Check	OK	23:00	

Data Download Date/Time: 06/26/2019 12:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,CHARLES,S
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: WILSON,PERRY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: LEDBETTER JOHN

DATE/TIME OF TEST: 06/16/2019 02:42:33

Diagnostics Check	OK	02:35
Air Blank	0.000	02:36
Control Test	0.080	02:36
Air Blank	0.000	02:36
Subject Sample # 1	0.142	02:37
Breath Volume	1.289	02:37
Air Blank	0.000	02:38
Air Blank	0.000	02:39
Subject Sample # 2	0.150	02:41
Breath Volume	1.164	02:41
Air Blank	0.000	02:41
Control Test	0.080	02:41
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 06/26/2019 12:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARKLEY,JEFFREY,G
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: BARKLEY,JEFFREY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: VERGARA AYALA

DATE/TIME OF TEST: 06/16/2019 03:19:50

Diagnostics Check	OK	03:13
Air Blank	0.000	03:13
Control Test	0.080	03:14
Air Blank	0.000	03:14
Subject Sample # 1	0.187	03:15
Breath Volume	2.773	03:15
Air Blank	0.000	03:15
Air Blank	0.000	03:17
Subject Sample # 2	0.186	03:18
Breath Volume	2.667	03:18
Air Blank	0.000	03:18
Control Test	0.079	03:19
Air Blank	0.000	03:19
Diagnostics Check	OK	03:19

Data Download Date/Time: 06/26/2019 12:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: HOYT

DATE/TIME OF TEST: 06/18/2019 01:29:38

Diagnostics Check	OK	01:22
Air Blank	0.000	01:22
Control Test	0.081	01:23
Air Blank	0.000	01:23
Subject Sample # 1	0.154	01:24
Breath Volume	4.117	01:24
Air Blank	0.000	01:25
Air Blank	0.000	01:27
Subject Sample # 2	0.162	01:28
Breath Volume	2.679	01:28
Air Blank	0.000	01:28
Control Test	0.079	01:29
Air Blank	0.000	01:29
Diagnostics Check	OK	01:29

Data Download Date/Time: 06/26/2019 12:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSH,DANIEL,C
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: FRAZIN,SHAWN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: LAMSON

DATE/TIME OF TEST: 06/18/2019 23:35:36

Diagnostics Check	OK	23:28
Air Blank	0.000	23:29
Control Test	0.081	23:29
Air Blank	0.000	23:29
Subject Sample # 1	0.223	23:31
Breath Volume	2.156	23:31
Air Blank	0.000	23:31
Air Blank	0.000	23:33
Subject Sample # 2	0.223	23:34
Breath Volume	2.683	23:34
Air Blank	0.000	23:34
Control Test	0.078	23:35
Air Blank	0.000	23:35
Diagnostics Check	OK	23:35

Data Download Date/Time: 06/26/2019 12:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: SHIPPY,DYLAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:47

SUBJECT'S LAST NAME: EMRAH

DATE/TIME OF TEST: 06/21/2019 22:14:42

Diagnostics Check	OK	22:07
Air Blank	0.000	22:08
Control Test	0.081	22:08
Air Blank	0.000	22:08
Subject Sample # 1	0.157	22:10
Breath Volume	1.683	22:10
Air Blank	0.000	22:10
Air Blank	0.000	22:12
Subject Sample # 2	0.153	22:13
Breath Volume	1.437	22:13
Air Blank	0.000	22:13
Control Test	0.079	22:14
Air Blank	0.000	22:14
Diagnostics Check	OK	22:14

Data Download Date/Time: 06/26/2019 12:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: POWERS

DATE/TIME OF TEST: 06/22/2019 03:45:25

Diagnostics Check	OK	03:39
Air Blank	0.000	03:39
Control Test	0.080	03:40
Air Blank	0.000	03:40
Subject Sample # 1	0.015	03:41
Breath Volume	1.394	03:41
Air Blank	0.000	03:41
Air Blank	0.000	03:43
Subject Sample # 2	0.013	03:44
Breath Volume	1.949	03:44
Air Blank	0.000	03:44
Control Test	0.079	03:44
Air Blank	0.000	03:45
Diagnostics Check	OK	03:45

Data Download Date/Time: 06/26/2019 12:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSH,DANIEL,C
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: RUSH,DANIEL
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: ARTZNER

DATE/TIME OF TEST: 06/23/2019 04:45:07

Diagnostics Check	OK	04:38
Air Blank	0.000	04:39
Control Test	0.080	04:39
Air Blank	0.000	04:39
Subject Sample # 1	0.086	04:40
Breath Volume	4.023	04:40
Air Blank	0.000	04:41
Air Blank	0.000	04:42
Subject Sample # 2	0.084	04:43
Breath Volume	4.644	04:43
Air Blank	0.000	04:44
Control Test	0.078	04:44
Air Blank	0.000	04:45
Diagnostics Check	OK	04:45

Data Download Date/Time: 06/26/2019 12:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOOS,SAVANNAH,K
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: CINTRON,M
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:29

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 06/23/2019 17:56:32

Diagnostics Check	OK	17:50	
Air Blank	0.000	17:50	
Control Test	0.080	17:51	
Air Blank	0.000	17:51	
Subject Sample # 1	0.249	17:52	
Breath Volume	1.972	17:52	
Air Blank	0.000	17:52	
Air Blank	0.000	17:54	
Subject Sample # 2		17:55	Subject Test Refused
Breath Volume	0.000	17:55	
Air Blank	0.000	17:55	
Control Test	0.080	17:55	
Air Blank	0.000	17:56	
Diagnostics Check	OK	17:56	

Data Download Date/Time: 06/26/2019 12:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FILLAK, MICHAEL, O
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: LEONARD, STEPHEN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: VONA

DATE/TIME OF TEST: 06/24/2019 20:59:23

Diagnostics Check	OK	20:51
Air Blank	0.000	20:52
Control Test	0.080	20:52
Air Blank	0.000	20:52
Subject Sample # 1	0.000	20:54
Breath Volume	3.375	20:54
Air Blank	0.000	20:54
Air Blank	0.000	20:56
Subject Sample # 2	0.000	20:58
Breath Volume	3.175	20:58
Air Blank	0.000	20:58
Control Test	0.080	20:58
Air Blank	0.000	20:59
Diagnostics Check	OK	20:59

Data Download Date/Time: 06/26/2019 12:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAMOUR,ZACHARY,T
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LESPERANCE,DANIEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: TYSON

DATE/TIME OF TEST: 06/14/2019 01:19:13

Diagnostics Check	OK	01:13	
Air Blank	0.000	01:14	
Control Test	0.078	01:14	
Air Blank	0.000	01:15	
Subject Sample	# 1	01:17	Subject Test Refused
Breath Volume	0.000	01:17	
Air Blank	0.000	01:18	
Control Test	0.078	01:18	
Air Blank	0.000	01:19	
Diagnostics Check	OK	01:19	

Data Download Date/Time: 07/11/2019 10:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,ALBERTO,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: GUZMAN,GUILLERMO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 05/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 620051
DRY GAS EXPIRY DATE: 11/05/2019
VIOLATION CODE: DUI
ARREST TIME: 13:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:03

SUBJECT'S LAST NAME: ANDREEN

DATE/TIME OF TEST: 06/23/2019 14:34:15

Diagnostics Check	OK	14:31	
Air Blank	0.000	14:31	
Control Test	0.080	14:32	
Air Blank	0.000	14:32	
Subject Sample	# 1	14:32	Subject Test Refused
Breath Volume	0.000	14:32	
Air Blank	0.000	14:33	
Control Test	0.079	14:33	
Air Blank	0.000	14:34	
Diagnostics Check	OK	14:34	

Data Download Date/Time: 06/27/2019 16:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CARRIEL,PATRICIA,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: CARRIEL,PATRICIA

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 05/31/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 620051

DRY GAS EXPIRY DATE: 11/05/2019

VIOLATION CODE: DUI

ARREST TIME: 16:46

GENDER: Female

OBSERVATION PERIOD BEGAN: 16:46

SUBJECT'S LAST NAME: WILCOX

DATE/TIME OF TEST: 06/23/2019 17:23:25

Diagnostics Check	OK	17:19	
Air Blank	0.000	17:19	
Control Test	0.081	17:20	
Air Blank	0.000	17:20	
Subject Sample	# 1	17:22	Sequence Aborted
Breath Volume	0.000	17:22	
Air Blank	0.000	17:23	

Data Download Date/Time: 06/27/2019 16:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOMEZ,JOSEPH,M
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: RIOS,R
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 05/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 620051
DRY GAS EXPIRY DATE: 11/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: GAVIRIA

DATE/TIME OF TEST: 06/23/2019 23:14:05

Diagnostics Check	OK	23:11	
Air Blank	0.000	23:11	
Control Test	0.080	23:11	
Air Blank	0.000	23:12	
Subject Sample	# 1	23:12	Subject Test Refused
Breath Volume	0.000	23:12	
Air Blank	0.000	23:13	
Control Test	0.080	23:13	
Air Blank	0.000	23:13	
Diagnostics Check	OK	23:14	

Data Download Date/Time: 06/27/2019 16:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASH,JUSTIN,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: TRUSCIO,JESSE
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 05/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 08:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:30

SUBJECT'S LAST NAME: PRICE

DATE/TIME OF TEST: 06/01/2019 09:56:56

Diagnostics Check	OK	09:52	
Air Blank	0.000	09:53	
Control Test	0.082	09:53	
Air Blank	0.000	09:54	
Subject Sample	# 1	09:55	Subject Test Refused
Breath Volume	0.000	09:55	
Air Blank	0.000	09:56	
Control Test	0.082	09:56	
Air Blank	0.000	09:56	
Diagnostics Check	OK	09:56	

Data Download Date/Time: 06/26/2019 13:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRESTON,MIKAYLA,D
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: TRUSCIO,JESSE
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 05/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 12:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:30

SUBJECT'S LAST NAME: NOURSE

DATE/TIME OF TEST: 06/02/2019 13:59:58

Diagnostics Check	OK	13:53
Air Blank	0.000	13:53
Control Test	0.082	13:53
Air Blank	0.000	13:54
Subject Sample # 1	0.112	13:55
Breath Volume	1.925	13:55
Air Blank	0.000	13:55
Air Blank	0.000	13:57
Subject Sample # 2	0.111	13:58
Breath Volume	2.156	13:58
Air Blank	0.000	13:59
Control Test	0.081	13:59
Air Blank	0.000	13:59
Diagnostics Check	OK	13:59

Data Download Date/Time: 06/26/2019 13:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRESTON,MIKAYLA,D
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: BRIGGS,JASON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 05/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 14:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:06

SUBJECT'S LAST NAME: FLOREZ

DATE/TIME OF TEST: 06/02/2019 15:31:03

Diagnostics Check	OK	15:28	
Air Blank	0.000	15:28	
Control Test	0.081	15:28	
Air Blank	0.000	15:29	
Subject Sample	# 1	15:29	Subject Test Refused
Breath Volume	0.000	15:29	
Air Blank	0.000	15:30	
Control Test	0.081	15:30	
Air Blank	0.000	15:30	
Diagnostics Check	OK	15:31	

Data Download Date/Time: 06/26/2019 13:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASH,JUSTIN,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: LAY,STEVEN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 05/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 12:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:10

SUBJECT'S LAST NAME: GARCIA HERNANDEZ

DATE/TIME OF TEST: 06/05/2019 13:43:28

Diagnostics Check	OK	13:37
Air Blank	0.000	13:37
Control Test	0.081	13:37
Air Blank	0.000	13:38
Subject Sample # 1	0.268	13:38
Breath Volume	2.207	13:38
Air Blank	0.000	13:39
Air Blank	0.000	13:41
Subject Sample # 2	0.262	13:41
Breath Volume	1.957	13:41
Air Blank	0.000	13:42
Control Test	0.081	13:42
Air Blank	0.000	13:43
Diagnostics Check	OK	13:43

Data Download Date/Time: 06/26/2019 13:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONADIE,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: BONADIE,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/29/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 17:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:06

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 06/08/2019 19:30:10

Diagnostics Check	OK	19:26	
Air Blank	0.000	19:26	
Control Test	0.082	19:26	
Air Blank	0.000	19:27	
Subject Sample	# 1	19:28	Subject Test Refused
Breath Volume	0.000	19:28	
Air Blank	0.000	19:29	
Control Test	0.081	19:29	
Air Blank	0.000	19:30	
Diagnostics Check	OK	19:30	

Data Download Date/Time: 06/26/2019 13:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,SEAN,C
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ANDERSON,JUSTIN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 05/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: WALTHOUR

DATE/TIME OF TEST: 06/13/2019 03:04:37

Diagnostics Check		OK	02:58
Air Blank		0.000	02:58
Control Test		0.082	02:58
Air Blank		0.000	02:59
Subject Sample	# 1	0.227	03:00
Breath Volume		2.718	03:00
Air Blank		0.000	03:00
Air Blank		0.000	03:02
Subject Sample	# 2	0.221	03:03
Breath Volume		2.687	03:03
Air Blank		0.000	03:03
Control Test		0.081	03:04
Air Blank		0.000	03:04
Diagnostics Check		OK	03:04

Data Download Date/Time: 06/26/2019 13:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHORTER,DENNIS,F
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: SHORTER,DENNIS
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 15:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: H452544626880

DATE/TIME OF TEST: 06/14/2019 21:19:16

Diagnostics Check	OK	21:06	
Air Blank	0.000	21:06	
Control Test	0.081	21:06	
Air Blank	0.000	21:07	
Subject Sample	# 1 0.206	21:09	
Breath Volume	1.113	21:09	
Air Blank	0.000	21:09	
Air Blank	0.000	21:11	
Subject Sample	# 2 0.183	21:14	Volume Not Met
Breath Volume	0.976	21:14	
Air Blank	0.000	21:15	
Air Blank	0.000	21:17	
Subject Sample	# 3 0.186	21:17	
Breath Volume	1.750	21:17	
Air Blank	0.000	21:18	
Control Test	0.076	21:18	
Air Blank	0.000	21:19	
Diagnostics Check	OK	21:19	

Data Download Date/Time: 06/26/2019 13:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRERA,ALEXANDER,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: CARSON,WILLIE
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 05/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: PONKEY

DATE/TIME OF TEST: 06/15/2019 04:27:30

Diagnostics Check	OK	04:20
Air Blank	0.000	04:21
Control Test	0.080	04:21
Air Blank	0.000	04:22
Subject Sample # 1	0.147	04:22
Breath Volume	2.015	04:22
Air Blank	0.000	04:23
Air Blank	0.000	04:25
Subject Sample # 2	0.148	04:25
Breath Volume	2.593	04:25
Air Blank	0.000	04:26
Control Test	0.081	04:26
Air Blank	0.000	04:27
Diagnostics Check	OK	04:27

Data Download Date/Time: 06/26/2019 13:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONADIE,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: BONADIE,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 11:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:19

SUBJECT'S LAST NAME: TERUEL

DATE/TIME OF TEST: 06/17/2019 12:47:02

Diagnostics Check	OK	12:40
Air Blank	0.000	12:40
Control Test	0.082	12:40
Air Blank	0.000	12:41
Subject Sample # 1	0.000	12:42
Breath Volume	2.582	12:42
Air Blank	0.000	12:42
Air Blank	0.000	12:44
Subject Sample # 2	0.000	12:45
Breath Volume	1.757	12:45
Air Blank	0.000	12:46
Control Test	0.081	12:46
Air Blank	0.000	12:46
Diagnostics Check	OK	12:47

Data Download Date/Time: 06/26/2019 13:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BALLOUN,SAMUEL,C
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: BALLOUN,SAMUEL
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 05/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 13:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:05

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 06/17/2019 14:35:42

Diagnostics Check	OK	14:29
Air Blank	0.000	14:29
Control Test	0.082	14:30
Air Blank	0.000	14:30
Subject Sample # 1	0.082	14:31
Breath Volume	2.359	14:31
Air Blank	0.000	14:31
Air Blank	0.000	14:33
Subject Sample # 2	0.080	14:34
Breath Volume	2.148	14:34
Air Blank	0.000	14:34
Control Test	0.081	14:35
Air Blank	0.000	14:35
Diagnostics Check	OK	14:35

Data Download Date/Time: 06/26/2019 13:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,A
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 05/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 11:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:00

SUBJECT'S LAST NAME: BALESTRA

DATE/TIME OF TEST: 06/20/2019 12:41:27

Diagnostics Check	OK	12:35
Air Blank	0.000	12:35
Control Test	0.082	12:35
Air Blank	0.000	12:36
Subject Sample # 1	0.262	12:36
Breath Volume	2.386	12:36
Air Blank	0.000	12:37
Air Blank	0.000	12:39
Subject Sample # 2	0.247	12:39
Breath Volume	2.222	12:39
Air Blank	0.000	12:40
Control Test	0.082	12:40
Air Blank	0.000	12:41
Diagnostics Check	OK	12:41

Data Download Date/Time: 06/26/2019 13:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMS,WALKER,L
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: BEATTIE,ISALAH
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 15:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:35

SUBJECT'S LAST NAME: HEINRICHS

DATE/TIME OF TEST: 06/28/2019 16:02:17

Diagnostics Check	OK	15:56
Air Blank	0.000	15:56
Control Test	0.082	15:56
Air Blank	0.000	15:57
Subject Sample # 1	0.235	15:57
Breath Volume	1.804	15:57
Air Blank	0.000	15:58
Air Blank	0.000	16:00
Subject Sample # 2	0.233	16:00
Breath Volume	1.500	16:00
Air Blank	0.000	16:01
Control Test	0.081	16:01
Air Blank	0.000	16:02
Diagnostics Check	OK	16:02

Data Download Date/Time: 07/16/2019 14:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001656

INSTRUMENT REGISTERED TO: NORTH MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PINEDA,MIGUEL,A
BREATH TEST OPERATOR AGENCY: NORTH MIAMI PD
ARRESTING OFFICER: FANOT,J
ARRESTING OFFICER AGENCY: NORTH MIAMI PD
LAST AGENCY INSPECTION: 05/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34418080A3
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:07

SUBJECT'S LAST NAME: CHARLES

DATE/TIME OF TEST: 06/09/2019 06:35:01

Diagnostics Check	OK	06:31	
Air Blank	0.000	06:32	
Control Test	0.082	06:32	
Air Blank	0.000	06:33	
Subject Sample	# 1	06:33	Subject Test Refused
Breath Volume	0.000	06:33	
Air Blank	0.000	06:34	
Control Test	0.082	06:34	
Air Blank	0.000	06:34	
Diagnostics Check	OK	06:35	

Data Download Date/Time: 06/21/2019 15:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001656

INSTRUMENT REGISTERED TO: NORTH MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PINEDA,MIGUEL,A
BREATH TEST OPERATOR AGENCY: NORTH MIAMI PD
ARRESTING OFFICER: MONTINA,N
ARRESTING OFFICER AGENCY: NORTH MIAMI PD
LAST AGENCY INSPECTION: 05/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34418080A3
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: SAID

DATE/TIME OF TEST: 06/09/2019 22:46:50

Diagnostics Check	OK	22:43	
Air Blank	0.000	22:44	
Control Test	0.081	22:44	
Air Blank	0.000	22:45	
Subject Sample	# 1	22:45	Subject Test Refused
Breath Volume	0.000	22:45	
Air Blank	0.000	22:45	
Control Test	0.081	22:46	
Air Blank	0.000	22:46	
Diagnostics Check	OK	22:46	

Data Download Date/Time: 06/21/2019 15:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001657

INSTRUMENT REGISTERED TO: MICCOSUKEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROCA,LUIS,C
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: DOMINGUEZ,HORACIO
ARRESTING OFFICER AGENCY: MICCOSUKEE PD
LAST AGENCY INSPECTION: 05/01/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 34917080A2
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: HERNANDEZ-GARCIA

DATE/TIME OF TEST: 06/09/2019 01:18:17

Diagnostics Check		OK	01:11
Air Blank		0.000	01:12
Control Test		0.077	01:12
Air Blank		0.000	01:12
Subject Sample	# 1	0.164	01:13
Breath Volume		2.027	01:13
Air Blank		0.000	01:14
Air Blank		0.000	01:16
Subject Sample	# 2	0.164	01:16
Breath Volume		2.421	01:16
Air Blank		0.000	01:17
Control Test		0.077	01:17
Air Blank		0.000	01:18
Diagnostics Check		OK	01:18

Data Download Date/Time: 06/25/2019 14:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001657

INSTRUMENT REGISTERED TO: MICCOSUKEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROMERO,OMAR,E
BREATH TEST OPERATOR AGENCY: MICCOSUKEE PD
ARRESTING OFFICER: DOMINGUEZ,HORACIO
ARRESTING OFFICER AGENCY: MICCOSUKEE PD
LAST AGENCY INSPECTION: 06/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A2
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: REYMUNDO

DATE/TIME OF TEST: 06/28/2019 20:24:18

Diagnostics Check	OK	20:16
Air Blank	0.000	20:16
Control Test	0.077	20:17
Air Blank	0.000	20:17
Subject Sample # 1	0.129	20:19
Breath Volume	2.648	20:19
Air Blank	0.000	20:19
Air Blank	0.000	20:21
Subject Sample # 2	0.135	20:22
Breath Volume	3.722	20:22
Air Blank	0.000	20:23
Control Test	0.076	20:23
Air Blank	0.000	20:24
Diagnostics Check	OK	20:24

Data Download Date/Time: 08/01/2019 09:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,DANIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: BRONSON,MILTON
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 05/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 06/01/2019 01:47:39

Diagnostics Check	OK	01:41
Air Blank	0.000	01:41
Control Test	0.079	01:42
Air Blank	0.000	01:42
Subject Sample # 1	0.061	01:42
Breath Volume	2.136	01:42
Air Blank	0.000	01:43
Air Blank	0.000	01:45
Subject Sample # 2	0.060	01:46
Breath Volume	2.183	01:46
Air Blank	0.000	01:46
Control Test	0.078	01:47
Air Blank	0.000	01:47
Diagnostics Check	OK	01:47

Data Download Date/Time: 06/11/2019 21:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAYERS,MICHELLE,H
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: SAYERS,MICHELLE
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: RIVERA-LORENZANA

DATE/TIME OF TEST: 06/01/2019 22:58:07

Diagnostics Check	OK	22:52
Air Blank	0.000	22:52
Control Test	0.078	22:52
Air Blank	0.000	22:53
Subject Sample # 1	0.000	22:53
Breath Volume	3.710	22:53
Air Blank	0.000	22:54
Air Blank	0.000	22:56
Subject Sample # 2	0.000	22:56
Breath Volume	3.902	22:56
Air Blank	0.000	22:57
Control Test	0.077	22:57
Air Blank	0.000	22:58
Diagnostics Check	OK	22:58

Data Download Date/Time: 06/11/2019 21:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AKINS,JUSTIN,M
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LOPEZ,MICHALL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: FLEURIDOR

DATE/TIME OF TEST: 06/02/2019 03:17:24

Diagnostics Check	OK	03:11
Air Blank	0.000	03:11
Control Test	0.079	03:11
Air Blank	0.000	03:12
Subject Sample # 1	0.053	03:13
Breath Volume	1.343	03:13
Air Blank	0.000	03:13
Air Blank	0.000	03:15
Subject Sample # 2	0.055	03:15
Breath Volume	1.890	03:15
Air Blank	0.000	03:16
Control Test	0.079	03:16
Air Blank	0.000	03:17
Diagnostics Check	OK	03:17

Data Download Date/Time: 06/11/2019 21:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PYE,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: PYE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: ESTREMERA-MOLINA

DATE/TIME OF TEST: 06/04/2019 01:50:03

Diagnostics Check	OK	01:43
Air Blank	0.000	01:43
Control Test	0.080	01:43
Air Blank	0.000	01:44
Subject Sample # 1	0.000	01:45
Breath Volume	2.210	01:45
Air Blank	0.000	01:46
Air Blank	0.000	01:48
Subject Sample # 2	0.000	01:48
Breath Volume	2.054	01:48
Air Blank	0.000	01:49
Control Test	0.079	01:49
Air Blank	0.000	01:49
Diagnostics Check	OK	01:50

Data Download Date/Time: 06/11/2019 21:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAYERS,MICHELLE,H
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: IRIZARRY,MIGUEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: MARQUEZ

DATE/TIME OF TEST: 06/06/2019 05:41:04

Diagnostics Check	OK	05:32
Air Blank	0.000	05:32
Control Test	0.080	05:33
Air Blank	0.000	05:33
Subject Sample # 1	0.195	05:36
Breath Volume	1.273	05:36
Air Blank	0.000	05:37
Air Blank	0.000	05:38
Subject Sample # 2	0.202	05:39
Breath Volume	1.957	05:39
Air Blank	0.000	05:40
Control Test	0.078	05:40
Air Blank	0.000	05:40
Diagnostics Check	OK	05:41

Data Download Date/Time: 06/11/2019 21:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAEZ,BAYRON,L
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: ANTHONY,COREY
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 05/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: MATIAS ACEVEDO

DATE/TIME OF TEST: 06/08/2019 23:52:24

Diagnostics Check	OK	23:46
Air Blank	0.000	23:46
Control Test	0.080	23:46
Air Blank	0.000	23:47
Subject Sample # 1	0.113	23:47
Breath Volume	2.828	23:47
Air Blank	0.000	23:48
Air Blank	0.000	23:50
Subject Sample # 2	0.101	23:50
Breath Volume	1.156	23:50
Air Blank	0.000	23:51
Control Test	0.078	23:51
Air Blank	0.000	23:52
Diagnostics Check	OK	23:52

Data Download Date/Time: 06/11/2019 21:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PYE,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: PYE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: BRESCIA

DATE/TIME OF TEST: 06/09/2019 04:10:10

Diagnostics Check	OK	04:02
Air Blank	0.000	04:02
Control Test	0.079	04:02
Air Blank	0.000	04:03
Subject Sample # 1	0.119	04:04
Breath Volume	1.480	04:04
Air Blank	0.000	04:05
Air Blank	0.000	04:06
Subject Sample # 2	0.113	04:08
Breath Volume	1.687	04:08
Air Blank	0.000	04:09
Control Test	0.077	04:09
Air Blank	0.000	04:10
Diagnostics Check	OK	04:10

Data Download Date/Time: 06/11/2019 21:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PYE,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: PYE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: DIAZ DIAZ

DATE/TIME OF TEST: 06/11/2019 05:05:47

Diagnostics Check	OK	04:59
Air Blank	0.000	04:59
Control Test	0.080	04:59
Air Blank	0.000	05:00
Subject Sample # 1	0.061	05:00
Breath Volume	2.078	05:00
Air Blank	0.000	05:01
Air Blank	0.000	05:03
Subject Sample # 2	0.061	05:04
Breath Volume	2.117	05:04
Air Blank	0.000	05:04
Control Test	0.078	05:05
Air Blank	0.000	05:05
Diagnostics Check	OK	05:05

Data Download Date/Time: 06/11/2019 21:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAYERS,MICHELLE,H
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: SAYERS,MICHELLE
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 06/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 06/12/2019 03:11:42

Diagnostics Check	OK	03:05
Air Blank	0.000	03:05
Control Test	0.079	03:06
Air Blank	0.000	03:06
Subject Sample # 1	0.000	03:07
Breath Volume	1.808	03:07
Air Blank	0.000	03:07
Air Blank	0.000	03:09
Subject Sample # 2	0.000	03:10
Breath Volume	1.578	03:10
Air Blank	0.000	03:10
Control Test	0.078	03:11
Air Blank	0.000	03:11
Diagnostics Check	OK	03:11

Data Download Date/Time: 07/02/2019 03:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCONNELL,CHARLES,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: MCCONNELL,CHARLES
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 06/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:33

SUBJECT'S LAST NAME: TOLEDO

DATE/TIME OF TEST: 06/12/2019 17:59:31

Diagnostics Check	OK	17:53
Air Blank	0.000	17:53
Control Test	0.079	17:54
Air Blank	0.000	17:54
Subject Sample # 1	0.000	17:55
Breath Volume	1.480	17:55
Air Blank	0.000	17:55
Air Blank	0.000	17:57
Subject Sample # 2	0.000	17:58
Breath Volume	1.226	17:58
Air Blank	0.000	17:58
Control Test	0.079	17:58
Air Blank	0.000	17:59
Diagnostics Check	OK	17:59

Data Download Date/Time: 07/02/2019 03:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCONNELL, CHARLES, E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CRAWFORD, DAVID
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 06/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: KELLAM

DATE/TIME OF TEST: 06/12/2019 23:30:58

Diagnostics Check	OK	23:24
Air Blank	0.000	23:25
Control Test	0.079	23:25
Air Blank	0.000	23:25
Subject Sample # 1	0.192	23:26
Breath Volume	2.316	23:26
Air Blank	0.000	23:27
Air Blank	0.000	23:28
Subject Sample # 2	0.198	23:29
Breath Volume	2.269	23:29
Air Blank	0.000	23:30
Control Test	0.078	23:30
Air Blank	0.000	23:30
Diagnostics Check	OK	23:30

Data Download Date/Time: 07/02/2019 03:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PYE,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: SAYERS,MICHELLE
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 06/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: NUNES

DATE/TIME OF TEST: 06/16/2019 05:46:47

Diagnostics Check	OK	05:38
Air Blank	0.000	05:38
Control Test	0.080	05:39
Air Blank	0.000	05:39
Subject Sample # 1	0.000	05:42
Breath Volume	1.261	05:42
Air Blank	0.000	05:43
Air Blank	0.000	05:45
Subject Sample # 2	0.000	05:45
Breath Volume	1.406	05:45
Air Blank	0.000	05:45
Control Test	0.078	05:46
Air Blank	0.000	05:46
Diagnostics Check	OK	05:46

Data Download Date/Time: 07/02/2019 03:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOUVRES,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: CRUZ,ASTRID
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 06/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:14

SUBJECT'S LAST NAME: RUIZ BALLEEN

DATE/TIME OF TEST: 06/16/2019 06:03:57

Diagnostics Check	OK	05:54
Air Blank	0.000	05:54
Control Test	0.080	05:55
Air Blank	0.000	05:55
Subject Sample # 1	0.075	05:58
Breath Volume	2.910	05:58
Air Blank	0.000	05:59
Air Blank	0.000	06:01
Subject Sample # 2	0.072	06:02
Breath Volume	2.269	06:02
Air Blank	0.000	06:03
Control Test	0.078	06:03
Air Blank	0.000	06:03
Diagnostics Check	OK	06:03

Data Download Date/Time: 07/02/2019 03:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAYERS,MICHELLE,H
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: SAYERS,MICHELLE
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 06/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: LOVADO-CEDIEL

DATE/TIME OF TEST: 06/17/2019 01:31:00

Diagnostics Check	OK	01:24
Air Blank	0.000	01:25
Control Test	0.081	01:25
Air Blank	0.000	01:26
Subject Sample # 1	0.241	01:26
Breath Volume	1.812	01:26
Air Blank	0.000	01:27
Air Blank	0.000	01:29
Subject Sample # 2	0.242	01:29
Breath Volume	2.394	01:29
Air Blank	0.000	01:30
Control Test	0.079	01:30
Air Blank	0.000	01:30
Diagnostics Check	OK	01:31

Data Download Date/Time: 07/02/2019 03:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,DANIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LYONS,DANIEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 06/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: OSORIO VARGAS

DATE/TIME OF TEST: 06/19/2019 03:44:53

Diagnostics Check	OK	03:38
Air Blank	0.000	03:39
Control Test	0.080	03:39
Air Blank	0.000	03:40
Subject Sample # 1	0.091	03:40
Breath Volume	1.210	03:40
Air Blank	0.000	03:41
Air Blank	0.000	03:42
Subject Sample # 2	0.091	03:43
Breath Volume	1.253	03:43
Air Blank	0.000	03:43
Control Test	0.079	03:44
Air Blank	0.000	03:44
Diagnostics Check	OK	03:44

Data Download Date/Time: 07/02/2019 03:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,DANIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CRAWFORD,DAVID
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 06/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: JIMENEZ JR

DATE/TIME OF TEST: 06/19/2019 04:14:31

Diagnostics Check	OK	04:07
Air Blank	0.000	04:07
Control Test	0.080	04:08
Air Blank	0.000	04:08
Subject Sample # 1	0.108	04:09
Breath Volume	1.394	04:09
Air Blank	0.000	04:10
Air Blank	0.000	04:12
Subject Sample # 2	0.116	04:12
Breath Volume	2.019	04:12
Air Blank	0.000	04:13
Control Test	0.078	04:13
Air Blank	0.000	04:14
Diagnostics Check	OK	04:14

Data Download Date/Time: 07/02/2019 03:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AKINS,JUSTIN,M
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: AKINS,JUSTIN
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 06/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: SUAREZ-MEDINA

DATE/TIME OF TEST: 06/20/2019 00:32:29

Diagnostics Check	OK	00:26
Air Blank	0.000	00:26
Control Test	0.081	00:27
Air Blank	0.000	00:27
Subject Sample # 1	0.000	00:28
Breath Volume	1.351	00:28
Air Blank	0.000	00:28
Air Blank	0.000	00:30
Subject Sample # 2	0.000	00:31
Breath Volume	1.152	00:31
Air Blank	0.000	00:31
Control Test	0.081	00:31
Air Blank	0.000	00:32
Diagnostics Check	OK	00:32

Data Download Date/Time: 07/02/2019 03:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOUVRES, MICHAEL, J
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: LANZO, MIGUEL
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 06/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: CRUCETA

DATE/TIME OF TEST: 06/21/2019 04:00:29

Diagnostics Check	OK	03:53
Air Blank	0.000	03:53
Control Test	0.080	03:54
Air Blank	0.000	03:54
Subject Sample # 1	0.089	03:55
Breath Volume	1.992	03:55
Air Blank	0.000	03:56
Air Blank	0.000	03:58
Subject Sample # 2	0.092	03:58
Breath Volume	1.621	03:58
Air Blank	0.000	03:59
Control Test	0.078	03:59
Air Blank	0.000	04:00
Diagnostics Check	OK	04:00

Data Download Date/Time: 07/02/2019 03:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,DANIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: SANTOS,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:27

SUBJECT'S LAST NAME: RUIZ PEREZ

DATE/TIME OF TEST: 06/23/2019 05:53:46

Diagnostics Check	OK	05:47
Air Blank	0.000	05:48
Control Test	0.080	05:48
Air Blank	0.000	05:48
Subject Sample # 1	0.088	05:49
Breath Volume	2.042	05:49
Air Blank	0.000	05:49
Air Blank	0.000	05:51
Subject Sample # 2	0.081	05:52
Breath Volume	1.382	05:52
Air Blank	0.000	05:52
Control Test	0.079	05:53
Air Blank	0.000	05:53
Diagnostics Check	OK	05:53

Data Download Date/Time: 07/02/2019 03:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PYE,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: PYE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 06/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: FEEZOR

DATE/TIME OF TEST: 06/26/2019 02:35:39

Diagnostics Check	OK	02:29
Air Blank	0.000	02:30
Control Test	0.080	02:30
Air Blank	0.000	02:30
Subject Sample # 1	0.000	02:31
Breath Volume	1.953	02:31
Air Blank	0.000	02:31
Air Blank	0.000	02:33
Subject Sample # 2	0.000	02:34
Breath Volume	2.574	02:34
Air Blank	0.000	02:34
Control Test	0.078	02:35
Air Blank	0.000	02:35
Diagnostics Check	OK	02:35

Data Download Date/Time: 07/02/2019 03:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MINERVINO,K,
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: MINERVINO,K
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 06/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 08:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:24

SUBJECT'S LAST NAME: GUILLAUME

DATE/TIME OF TEST: 06/26/2019 09:18:17

Diagnostics Check		OK	09:06	
Air Blank		0.000	09:06	
Control Test		0.079	09:07	
Air Blank		0.000	09:07	
Subject Sample	# 1	0.168	09:10	Volume Not Met
Breath Volume		0.386	09:10	
Air Blank		0.000	09:11	
Air Blank		0.000	09:13	
Subject Sample	# 2	0.174	09:16	Volume Not Met
Breath Volume		0.523	09:16	
Air Blank		0.000	09:17	
Control Test		0.076	09:17	
Air Blank		0.000	09:18	
Diagnostics Check		OK	09:18	

Data Download Date/Time: 07/02/2019 03:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTOS,JOSEPH,N
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: SANTOS,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:30

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 06/27/2019 06:58:12

Diagnostics Check	OK	06:51
Air Blank	0.000	06:52
Control Test	0.080	06:52
Air Blank	0.000	06:52
Subject Sample # 1	0.103	06:53
Breath Volume	1.648	06:53
Air Blank	0.000	06:54
Air Blank	0.000	06:56
Subject Sample # 2	0.099	06:56
Breath Volume	1.457	06:56
Air Blank	0.000	06:57
Control Test	0.079	06:57
Air Blank	0.000	06:58
Diagnostics Check	OK	06:58

Data Download Date/Time: 07/02/2019 03:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSELL,ROBERT,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: OLKA,MATTHEW
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 06/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:33

SUBJECT'S LAST NAME: CARABALLO

DATE/TIME OF TEST: 06/28/2019 06:06:07

Diagnostics Check		OK	05:56
Air Blank		0.000	05:56
Control Test		0.079	05:57
Air Blank		0.000	05:57
Subject Sample	# 1	0.224	05:58
Breath Volume		2.796	05:58
Air Blank		0.000	05:59
Air Blank		0.000	06:00
Subject Sample	# 2	0.193	06:01
Breath Volume		2.195	06:01
Air Blank		0.000	06:02
Air Blank		0.000	06:03
Subject Sample	# 3	0.200	06:04
Breath Volume		2.414	06:04
Air Blank		0.000	06:05
Control Test		0.076	06:05
Air Blank		0.000	06:06
Diagnostics Check		OK	06:06

Data Download Date/Time: 07/02/2019 03:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAYERS,MICHELLE,H
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: SAYERS,MICHELLE
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 06/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: QUIJANO-NIEVES

DATE/TIME OF TEST: 06/29/2019 02:28:16

Diagnostics Check	OK	02:21
Air Blank	0.000	02:22
Control Test	0.078	02:22
Air Blank	0.000	02:23
Subject Sample # 1	0.138	02:23
Breath Volume	1.625	02:23
Air Blank	0.000	02:24
Air Blank	0.000	02:25
Subject Sample # 2	0.139	02:26
Breath Volume	2.167	02:26
Air Blank	0.000	02:27
Control Test	0.077	02:27
Air Blank	0.000	02:28
Diagnostics Check	OK	02:28

Data Download Date/Time: 07/02/2019 03:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLLIER,THOMAS,L
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: COLLIER,THOMAS
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 06/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: NAPOLES-RAXACH

DATE/TIME OF TEST: 06/29/2019 03:52:16

Diagnostics Check	OK	03:45	
Air Blank	0.000	03:45	
Control Test	0.076	03:46	
Air Blank	0.000	03:46	
Subject Sample # 1	0.102	03:47	
Breath Volume	3.843	03:47	
Air Blank	0.000	03:48	
Air Blank	0.000	03:49	
Subject Sample # 2	0.097	03:50	
Breath Volume	2.750	03:50	
Air Blank	0.000	03:51	
Control Test	0.074	03:51	Control Outside Tolerance
Air Blank	0.000	03:52	

Data Download Date/Time: 07/02/2019 03:52