

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GODENZI,CAMERON,M
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: LAWRENCE,TYLER
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 07/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: CHAMPEON

DATE/TIME OF TEST: 07/03/2019 00:42:02

Diagnostics Check	OK	00:35
Air Blank	0.000	00:35
Control Test	0.079	00:35
Air Blank	0.000	00:36
Subject Sample # 1	0.191	00:37
Breath Volume	1.316	00:37
Air Blank	0.000	00:38
Air Blank	0.000	00:39
Subject Sample # 2	0.188	00:40
Breath Volume	2.328	00:40
Air Blank	0.000	00:41
Control Test	0.079	00:41
Air Blank	0.000	00:41
Diagnostics Check	OK	00:42

Data Download Date/Time: 09/03/2019 09:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LICKISS,BRANT,C
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: LICKISS,BRANT
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 07/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 07/05/2019 00:57:33

Diagnostics Check	OK	00:49
Air Blank	0.000	00:49
Control Test	0.080	00:50
Air Blank	0.000	00:50
Subject Sample # 1	0.018	00:53
Breath Volume	3.414	00:53
Air Blank	0.000	00:53
Air Blank	0.000	00:55
Subject Sample # 2	0.015	00:56
Breath Volume	3.003	00:56
Air Blank	0.000	00:56
Control Test	0.079	00:56
Air Blank	0.000	00:57
Diagnostics Check	OK	00:57

Data Download Date/Time: 09/03/2019 09:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,BRENT,M
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: NEALIS,JACOB
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 07/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: NORRIS

DATE/TIME OF TEST: 07/08/2019 04:03:01

Diagnostics Check		OK	03:56
Air Blank		0.000	03:57
Control Test		0.080	03:57
Air Blank		0.000	03:57
Subject Sample	# 1	0.166	03:58
Breath Volume		2.433	03:58
Air Blank		0.000	03:59
Air Blank		0.000	04:01
Subject Sample	# 2	0.168	04:01
Breath Volume		2.410	04:01
Air Blank		0.000	04:02
Control Test		0.080	04:02
Air Blank		0.000	04:02
Diagnostics Check		OK	04:03

Data Download Date/Time: 09/03/2019 09:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDGEComb,STEVEN,J
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: NEALIS,JACOB
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 07/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: PARODI

DATE/TIME OF TEST: 07/11/2019 00:08:58

Diagnostics Check	OK	00:02
Air Blank	0.000	00:03
Control Test	0.081	00:03
Air Blank	0.000	00:04
Subject Sample # 1	0.000	00:04
Breath Volume	1.523	00:04
Air Blank	0.000	00:05
Air Blank	0.000	00:07
Subject Sample # 2	0.000	00:07
Breath Volume	2.019	00:07
Air Blank	0.000	00:08
Control Test	0.080	00:08
Air Blank	0.000	00:08
Diagnostics Check	OK	00:08

Data Download Date/Time: 09/03/2019 09:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GODENZI,CAMERON,M
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: GODENZI,CAMERON
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 07/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: VERMILLION

DATE/TIME OF TEST: 07/28/2019 00:41:16

Diagnostics Check	OK	00:34
Air Blank	0.000	00:35
Control Test	0.081	00:35
Air Blank	0.000	00:36
Subject Sample # 1	0.122	00:36
Breath Volume	3.101	00:36
Air Blank	0.000	00:37
Air Blank	0.000	00:39
Subject Sample # 2	0.121	00:39
Breath Volume	3.347	00:39
Air Blank	0.000	00:40
Control Test	0.080	00:40
Air Blank	0.000	00:41
Diagnostics Check	OK	00:41

Data Download Date/Time: 09/03/2019 09:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'CONNELL,SHANNON,N
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: GHICA,VLAD
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 07/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: CZIRR

DATE/TIME OF TEST: 07/29/2019 00:13:31

Diagnostics Check	OK	00:04
Air Blank	0.000	00:04
Control Test	0.080	00:04
Air Blank	0.000	00:05
Subject Sample # 1	0.191	00:08
Breath Volume	1.453	00:08
Air Blank	0.000	00:09
Air Blank	0.000	00:11
Subject Sample # 2	0.201	00:11
Breath Volume	1.105	00:11
Air Blank	0.000	00:12
Control Test	0.081	00:12
Air Blank	0.000	00:13
Diagnostics Check	OK	00:13

Data Download Date/Time: 09/03/2019 09:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GODENZI,CAMERON,M
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: BROWN,TREVOR
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 07/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: NUNEZ

DATE/TIME OF TEST: 07/31/2019 02:28:32

Diagnostics Check	OK	02:20
Air Blank	0.000	02:21
Control Test	0.080	02:21
Air Blank	0.000	02:22
Subject Sample # 1	0.000	02:23
Breath Volume	1.554	02:23
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample # 2	0.000	02:27
Breath Volume	2.175	02:27
Air Blank	0.000	02:27
Control Test	0.081	02:27
Air Blank	0.000	02:28
Diagnostics Check	OK	02:28

Data Download Date/Time: 09/03/2019 09:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORREA,JONATHAN,E
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: CORREA,JONATHAN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 07/02/2019 01:38:15

Diagnostics Check	OK	01:32
Air Blank	0.000	01:32
Control Test	0.081	01:32
Air Blank	0.000	01:33
Subject Sample # 1	0.165	01:33
Breath Volume	2.078	01:33
Air Blank	0.000	01:34
Air Blank	0.000	01:36
Subject Sample # 2	0.164	01:36
Breath Volume	2.039	01:36
Air Blank	0.000	01:37
Control Test	0.080	01:37
Air Blank	0.000	01:38
Diagnostics Check	OK	01:38

Data Download Date/Time: 07/16/2019 14:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: HAMRICK

DATE/TIME OF TEST: 07/03/2019 03:13:08

Diagnostics Check	OK	03:10	
Air Blank	0.000	03:10	
Control Test	0.080	03:11	
Air Blank	0.000	03:11	
Subject Sample	# 1	03:11	Subject Test Refused
Breath Volume	0.000	03:11	
Air Blank	0.000	03:12	
Control Test	0.080	03:12	
Air Blank	0.000	03:13	
Diagnostics Check	OK	03:13	

Data Download Date/Time: 07/16/2019 14:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHIRSHIKOVA,ALINA,A
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: SIMMONS,KYLE
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 18:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:56

SUBJECT'S LAST NAME: MUNGAL

DATE/TIME OF TEST: 07/04/2019 19:33:42

Diagnostics Check	OK	19:27
Air Blank	0.000	19:27
Control Test	0.079	19:28
Air Blank	0.000	19:28
Subject Sample # 1	0.240	19:29
Breath Volume	4.101	19:29
Air Blank	0.000	19:29
Air Blank	0.000	19:31
Subject Sample # 2	0.239	19:32
Breath Volume	3.945	19:32
Air Blank	0.000	19:32
Control Test	0.078	19:33
Air Blank	0.000	19:33
Diagnostics Check	OK	19:33

Data Download Date/Time: 07/16/2019 14:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RADO,RYAN,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: RADO,RYAN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 22:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: LABORDE

DATE/TIME OF TEST: 07/04/2019 23:50:13

Diagnostics Check	OK	23:42
Air Blank	0.000	23:42
Control Test	0.080	23:42
Air Blank	0.000	23:43
Subject Sample # 1	0.148	23:45
Breath Volume	2.203	23:45
Air Blank	0.000	23:46
Air Blank	0.000	23:47
Subject Sample # 2	0.151	23:48
Breath Volume	2.265	23:48
Air Blank	0.000	23:49
Control Test	0.078	23:49
Air Blank	0.000	23:50
Diagnostics Check	OK	23:50

Data Download Date/Time: 07/16/2019 14:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORREA,JONATHAN,E
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: CORREA,JONATHAN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: CANON

DATE/TIME OF TEST: 07/05/2019 00:07:48

Diagnostics Check	OK	00:01
Air Blank	0.000	00:01
Control Test	0.080	00:01
Air Blank	0.000	00:02
Subject Sample # 1	0.186	00:03
Breath Volume	1.250	00:03
Air Blank	0.000	00:03
Air Blank	0.000	00:05
Subject Sample # 2	0.180	00:06
Breath Volume	1.296	00:06
Air Blank	0.000	00:06
Control Test	0.079	00:07
Air Blank	0.000	00:07
Diagnostics Check	OK	00:07

Data Download Date/Time: 07/16/2019 14:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: BYASSEE

DATE/TIME OF TEST: 07/11/2019 23:05:52

Diagnostics Check	OK	22:58
Air Blank	0.000	22:58
Control Test	0.080	22:59
Air Blank	0.000	22:59
Subject Sample # 1	0.140	23:01
Breath Volume	2.343	23:01
Air Blank	0.000	23:01
Air Blank	0.000	23:03
Subject Sample # 2	0.137	23:04
Breath Volume	3.761	23:04
Air Blank	0.000	23:04
Control Test	0.079	23:05
Air Blank	0.000	23:05
Diagnostics Check	OK	23:05

Data Download Date/Time: 07/16/2019 14:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORREA,JONATHAN,E
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: CORREA,JONATHAN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: STAFFORD

DATE/TIME OF TEST: 07/12/2019 22:24:44

Diagnostics Check	OK	22:18
Air Blank	0.000	22:19
Control Test	0.080	22:19
Air Blank	0.000	22:20
Subject Sample # 1	0.000	22:20
Breath Volume	1.398	22:20
Air Blank	0.000	22:21
Air Blank	0.000	22:23
Subject Sample # 2	0.000	22:23
Breath Volume	1.261	22:23
Air Blank	0.000	22:23
Control Test	0.080	22:24
Air Blank	0.000	22:24
Diagnostics Check	OK	22:24

Data Download Date/Time: 07/16/2019 14:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORD, MICHAEL, A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: GIANNOTTA, DOMINIC
ARRESTING OFFICER AGENCY: St. Augustine Beach
LAST AGENCY INSPECTION: 07/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 12:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:20

SUBJECT'S LAST NAME: GREENE

DATE/TIME OF TEST: 07/18/2019 13:49:57

Diagnostics Check	OK	13:43
Air Blank	0.000	13:43
Control Test	0.079	13:44
Air Blank	0.000	13:44
Subject Sample # 1	0.158	13:45
Breath Volume	3.710	13:45
Air Blank	0.000	13:46
Air Blank	0.000	13:47
Subject Sample # 2	0.156	13:48
Breath Volume	3.671	13:48
Air Blank	0.000	13:49
Control Test	0.078	13:49
Air Blank	0.000	13:49
Diagnostics Check	OK	13:49

Data Download Date/Time: 08/08/2019 12:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEORGE,KYLE,J
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HEDRICK,H
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 07/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 18:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:24

SUBJECT'S LAST NAME: DAMROW

DATE/TIME OF TEST: 07/18/2019 19:53:13

Diagnostics Check	OK	19:50	
Air Blank	0.000	19:50	
Control Test	0.078	19:51	
Air Blank	0.000	19:51	
Subject Sample	# 1	19:51	Subject Test Refused
Breath Volume	0.000	19:51	
Air Blank	0.000	19:52	
Control Test	0.079	19:52	
Air Blank	0.000	19:53	
Diagnostics Check	OK	19:53	

Data Download Date/Time: 08/08/2019 12:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DA SILVA,ANDRE,F
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: DA SILVA,ANDRE
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 07/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 17:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:01

SUBJECT'S LAST NAME: CONWELL

DATE/TIME OF TEST: 07/19/2019 18:29:19

Diagnostics Check		OK	18:22
Air Blank		0.000	18:23
Control Test		0.078	18:23
Air Blank		0.000	18:24
Subject Sample	# 1	0.201	18:24
Breath Volume		2.425	18:24
Air Blank		0.000	18:25
Air Blank		0.000	18:27
Subject Sample	# 2	0.205	18:27
Breath Volume		2.539	18:27
Air Blank		0.000	18:28
Control Test		0.079	18:28
Air Blank		0.000	18:29
Diagnostics Check		OK	18:29

Data Download Date/Time: 08/08/2019 12:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,MASON,E
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: THOMPSON,MASON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 07/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: HAVLINEK

DATE/TIME OF TEST: 07/20/2019 03:43:51

Diagnostics Check	OK	03:41	
Air Blank	0.000	03:41	
Control Test	0.079	03:41	
Air Blank	0.000	03:42	
Subject Sample	# 1	03:42	Subject Test Refused
Breath Volume	0.000	03:42	
Air Blank	0.000	03:42	
Control Test	0.078	03:43	
Air Blank	0.000	03:43	
Diagnostics Check	OK	03:43	

Data Download Date/Time: 08/08/2019 12:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,MASON,E
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: THOMPSON,MASON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 07/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: SEAVEY

DATE/TIME OF TEST: 07/20/2019 23:13:57

Diagnostics Check	OK	23:11	
Air Blank	0.000	23:11	
Control Test	0.078	23:11	
Air Blank	0.000	23:12	
Subject Sample	# 1	23:12	Subject Test Refused
Breath Volume	0.000	23:12	
Air Blank	0.000	23:13	
Control Test	0.079	23:13	
Air Blank	0.000	23:13	
Diagnostics Check	OK	23:13	

Data Download Date/Time: 08/08/2019 12:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 07/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: BENITEZ

DATE/TIME OF TEST: 07/25/2019 01:56:41

Diagnostics Check	OK	01:53	
Air Blank	0.000	01:54	
Control Test	0.079	01:54	
Air Blank	0.000	01:55	
Subject Sample	# 1	01:55	Subject Test Refused
Breath Volume	0.000	01:55	
Air Blank	0.000	01:55	
Control Test	0.079	01:56	
Air Blank	0.000	01:56	
Diagnostics Check	OK	01:56	

Data Download Date/Time: 08/08/2019 12:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 07/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 23:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: WANCHIK

DATE/TIME OF TEST: 07/26/2019 00:14:22

Diagnostics Check	OK	00:07
Air Blank	0.000	00:08
Control Test	0.079	00:08
Air Blank	0.000	00:09
Subject Sample # 1	0.054	00:09
Breath Volume	3.300	00:09
Air Blank	0.000	00:10
Air Blank	0.000	00:12
Subject Sample # 2	0.054	00:12
Breath Volume	3.183	00:12
Air Blank	0.000	00:13
Control Test	0.078	00:13
Air Blank	0.000	00:14
Diagnostics Check	OK	00:14

Data Download Date/Time: 08/08/2019 12:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENNESSEY, BRANDON, L
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: TRUSCIO, JESSE
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 07/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 16:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:23

SUBJECT'S LAST NAME: SUGGS

DATE/TIME OF TEST: 07/27/2019 18:15:54

Diagnostics Check	OK	18:09
Air Blank	0.000	18:09
Control Test	0.078	18:09
Air Blank	0.000	18:10
Subject Sample # 1	0.240	18:10
Breath Volume	2.769	18:10
Air Blank	0.000	18:11
Air Blank	0.000	18:13
Subject Sample # 2	0.224	18:14
Breath Volume	3.363	18:14
Air Blank	0.000	18:14
Control Test	0.078	18:15
Air Blank	0.000	18:15
Diagnostics Check	OK	18:15

Data Download Date/Time: 08/08/2019 12:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,STERLING,R
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: WEKONY,ELIJAH
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 06/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: MATHIS

DATE/TIME OF TEST: 07/05/2019 00:50:51

Diagnostics Check		OK	00:40
Air Blank		0.000	00:40
Control Test		0.082	00:41
Air Blank		0.000	00:41
Subject Sample	# 1	0.138	00:42
Breath Volume		1.710	00:42
Air Blank		0.000	00:42
Air Blank		0.000	00:44
Subject Sample	# 2	0.117	00:45
Breath Volume		1.750	00:45
Air Blank		0.000	00:46
Air Blank		0.000	00:48
Subject Sample	# 3	0.114	00:49
Breath Volume		1.832	00:49
Air Blank		0.000	00:49
Control Test		0.081	00:50
Air Blank		0.000	00:50
Diagnostics Check		OK	00:50

Data Download Date/Time: 07/22/2019 16:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,STERLING,R
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: BAZE,KAYLOR
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 06/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: MURILLO

DATE/TIME OF TEST: 07/05/2019 03:06:23

Diagnostics Check	OK	02:58
Air Blank	0.000	02:59
Control Test	0.082	02:59
Air Blank	0.000	03:00
Subject Sample # 1	0.048	03:01
Breath Volume	2.093	03:01
Air Blank	0.000	03:02
Air Blank	0.000	03:04
Subject Sample # 2	0.047	03:04
Breath Volume	2.601	03:04
Air Blank	0.000	03:05
Control Test	0.080	03:05
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

Data Download Date/Time: 07/22/2019 16:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNNE,JUSTIN,R
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: BODKIN,MICHAEL
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 06/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: CORHERN

DATE/TIME OF TEST: 07/06/2019 01:04:45

Diagnostics Check	OK	00:57
Air Blank	0.000	00:58
Control Test	0.082	00:58
Air Blank	0.000	00:59
Subject Sample # 1	0.000	01:00
Breath Volume	1.523	01:00
Air Blank	0.000	01:00
Air Blank	0.000	01:02
Subject Sample # 2	0.000	01:03
Breath Volume	1.570	01:03
Air Blank	0.000	01:03
Control Test	0.081	01:04
Air Blank	0.000	01:04
Diagnostics Check	OK	01:04

Data Download Date/Time: 07/22/2019 16:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNNE,JUSTIN,R
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: TORRES,LYDIA
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 06/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: PINEDA ULLOA

DATE/TIME OF TEST: 07/07/2019 02:08:11

Diagnostics Check	OK	02:02	
Air Blank	0.000	02:03	
Control Test	0.081	02:03	
Air Blank	0.000	02:04	
Subject Sample	# 1	02:06	Subject Test Refused
Breath Volume	0.253	02:06	
Air Blank	0.000	02:07	
Control Test	0.081	02:07	
Air Blank	0.000	02:08	
Diagnostics Check	OK	02:08	

Data Download Date/Time: 07/22/2019 16:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNNE,JUSTIN,R
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: DUNNE,JUSTIN
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 06/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: LARocca

DATE/TIME OF TEST: 07/09/2019 23:41:25

Diagnostics Check	OK	23:38	
Air Blank	0.000	23:39	
Control Test	0.082	23:39	
Air Blank	0.000	23:39	
Subject Sample	# 1	23:40	Subject Test Refused
Breath Volume	0.000	23:40	
Air Blank	0.000	23:40	
Control Test	0.081	23:40	
Air Blank	0.000	23:41	
Diagnostics Check	OK	23:41	

Data Download Date/Time: 07/22/2019 16:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNNE,JUSTIN,R
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: DUNNE,JUSTIN
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 06/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 22:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: MARTINI

DATE/TIME OF TEST: 07/10/2019 22:59:35

Diagnostics Check	OK	22:52
Air Blank	0.000	22:52
Control Test	0.081	22:52
Air Blank	0.000	22:53
Subject Sample # 1	0.000	22:54
Breath Volume	2.695	22:54
Air Blank	0.000	22:55
Air Blank	0.000	22:57
Subject Sample # 2	0.000	22:58
Breath Volume	2.203	22:58
Air Blank	0.000	22:58
Control Test	0.081	22:58
Air Blank	0.000	22:59
Diagnostics Check	OK	22:59

Data Download Date/Time: 07/22/2019 16:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,JAMES,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: QUINONES,BRIAN
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 06/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: SANDS

DATE/TIME OF TEST: 07/21/2019 00:28:17

Diagnostics Check		OK	00:20
Air Blank		0.000	00:20
Control Test		0.081	00:21
Air Blank		0.000	00:21
Subject Sample	# 1	0.144	00:23
Breath Volume		1.792	00:23
Air Blank		0.000	00:24
Air Blank		0.000	00:25
Subject Sample	# 2	0.139	00:26
Breath Volume		1.304	00:26
Air Blank		0.000	00:27
Control Test		0.079	00:27
Air Blank		0.000	00:28
Diagnostics Check		OK	00:28

Data Download Date/Time: 07/22/2019 16:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,JAMES,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: WARD,JAMES
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 06/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: REID

DATE/TIME OF TEST: 07/21/2019 03:28:46

Diagnostics Check		OK	03:21
Air Blank		0.000	03:22
Control Test		0.081	03:22
Air Blank		0.000	03:22
Subject Sample	# 1	0.238	03:23
Breath Volume		2.757	03:23
Air Blank		0.000	03:24
Air Blank		0.000	03:26
Subject Sample	# 2	0.225	03:27
Breath Volume		2.507	03:27
Air Blank		0.000	03:27
Control Test		0.080	03:28
Air Blank		0.000	03:28
Diagnostics Check		OK	03:28

Data Download Date/Time: 07/22/2019 16:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001136

INSTRUMENT REGISTERED TO: DELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,PAUL,A
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: LIZABET,EDUARD
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: RILEY

DATE/TIME OF TEST: 07/08/2019 22:54:36

Diagnostics Check	OK	22:51	
Air Blank	0.000	22:52	
Control Test	0.080	22:52	
Air Blank	0.000	22:53	
Subject Sample	# 1	22:53	Subject Test Refused
Breath Volume	0.000	22:53	
Air Blank	0.000	22:53	
Control Test	0.081	22:54	
Air Blank	0.000	22:54	
Diagnostics Check	OK	22:54	

Data Download Date/Time: 07/31/2019 16:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001136

INSTRUMENT REGISTERED TO: DELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDGEComb,STEVEN,J
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: STOLTZ,MICAH
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 04:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:02

SUBJECT'S LAST NAME: AGUIRRE

DATE/TIME OF TEST: 07/28/2019 05:31:02

Diagnostics Check	OK	05:24
Air Blank	0.000	05:25
Control Test	0.080	05:25
Air Blank	0.000	05:25
Subject Sample # 1	0.119	05:26
Breath Volume	1.351	05:26
Air Blank	0.000	05:27
Air Blank	0.000	05:29
Subject Sample # 2	0.119	05:29
Breath Volume	1.378	05:29
Air Blank	0.000	05:30
Control Test	0.080	05:30
Air Blank	0.000	05:30
Diagnostics Check	OK	05:31

Data Download Date/Time: 07/31/2019 16:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHEWNING,AUSTIN,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: CONWAY,PHILIP
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:35

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 07/01/2019 18:09:01

Diagnostics Check	OK	18:06	
Air Blank	0.000	18:06	
Control Test	0.079	18:07	
Air Blank	0.000	18:07	
Subject Sample	# 1	18:07	Subject Test Refused
Breath Volume	0.000	18:07	
Air Blank	0.000	18:08	
Control Test	0.080	18:08	
Air Blank	0.000	18:08	
Diagnostics Check	OK	18:09	

Data Download Date/Time: 07/28/2019 10:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROUCH,JOSHUA,L
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: GOSSETT,ADAM
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: MAXCY

DATE/TIME OF TEST: 07/02/2019 21:58:16

Diagnostics Check		OK	21:51
Air Blank		0.000	21:51
Control Test		0.080	21:52
Air Blank		0.000	21:52
Subject Sample	# 1	0.333	21:53
Breath Volume		1.945	21:53
Air Blank		0.000	21:53
Air Blank		0.000	21:55
Subject Sample	# 2	0.344	21:56
Breath Volume		1.832	21:56
Air Blank		0.000	21:57
Control Test		0.080	21:57
Air Blank		0.000	21:58
Diagnostics Check		OK	21:58

Data Download Date/Time: 07/28/2019 10:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROUCH,JOSHUA,L
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: MALTA,D
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/03/2019 03:34:51

Diagnostics Check	OK	03:31	
Air Blank	0.000	03:32	
Control Test	0.081	03:32	
Air Blank	0.000	03:32	
Subject Sample	# 1	03:33	Subject Test Refused
Breath Volume	0.000	03:33	
Air Blank	0.000	03:33	
Control Test	0.081	03:34	
Air Blank	0.000	03:34	
Diagnostics Check	OK	03:34	

Data Download Date/Time: 07/28/2019 10:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,CARL,L
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: OBARR,JOSEPH
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:32

SUBJECT'S LAST NAME: MOREA

DATE/TIME OF TEST: 07/03/2019 18:44:19

Diagnostics Check	OK	18:37
Air Blank	0.000	18:37
Control Test	0.081	18:38
Air Blank	0.000	18:38
Subject Sample # 1	0.000	18:39
Breath Volume	1.281	18:39
Air Blank	0.000	18:40
Air Blank	0.000	18:42
Subject Sample # 2	0.000	18:42
Breath Volume	1.683	18:42
Air Blank	0.000	18:43
Control Test	0.080	18:43
Air Blank	0.000	18:44
Diagnostics Check	OK	18:44

Data Download Date/Time: 07/28/2019 10:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ST JOHN,J,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: DENKER,B
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: LINDSEY

DATE/TIME OF TEST: 07/04/2019 04:28:30

Diagnostics Check	OK	04:21
Air Blank	0.000	04:22
Control Test	0.081	04:22
Air Blank	0.000	04:23
Subject Sample # 1	0.194	04:23
Breath Volume	2.847	04:23
Air Blank	0.000	04:24
Air Blank	0.000	04:25
Subject Sample # 2	0.185	04:26
Breath Volume	2.250	04:26
Air Blank	0.000	04:27
Control Test	0.079	04:27
Air Blank	0.000	04:28
Diagnostics Check	OK	04:28

Data Download Date/Time: 07/28/2019 10:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONG, JOSHUA, A
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: GOSSETT, ADAM
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: KRODEL

DATE/TIME OF TEST: 07/04/2019 23:28:42

Diagnostics Check	OK	23:22
Air Blank	0.000	23:22
Control Test	0.081	23:22
Air Blank	0.000	23:23
Subject Sample # 1	0.133	23:24
Breath Volume	1.804	23:24
Air Blank	0.000	23:24
Air Blank	0.000	23:26
Subject Sample # 2	0.120	23:27
Breath Volume	1.855	23:27
Air Blank	0.000	23:27
Control Test	0.080	23:28
Air Blank	0.000	23:28
Diagnostics Check	OK	23:28

Data Download Date/Time: 07/28/2019 10:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEAVER,DANIEL,
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: WEAVER,DANIEL
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: MAYO

DATE/TIME OF TEST: 07/06/2019 23:45:03

Diagnostics Check	OK	23:38
Air Blank	0.000	23:39
Control Test	0.080	23:39
Air Blank	0.000	23:40
Subject Sample # 1	0.132	23:40
Breath Volume	2.597	23:40
Air Blank	0.000	23:40
Air Blank	0.000	23:42
Subject Sample # 2	0.128	23:43
Breath Volume	1.160	23:43
Air Blank	0.000	23:44
Control Test	0.080	23:44
Air Blank	0.000	23:44
Diagnostics Check	OK	23:45

Data Download Date/Time: 07/28/2019 10:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAINEY,CRISTA,N
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: GOSSETT,ADAM
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 05:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:55

SUBJECT'S LAST NAME: ALLAN

DATE/TIME OF TEST: 07/07/2019 07:18:21

Diagnostics Check	OK	07:15	
Air Blank	0.000	07:15	
Control Test	0.080	07:16	
Air Blank	0.000	07:16	
Subject Sample	# 1	07:17	Subject Test Refused
Breath Volume	0.000	07:17	
Air Blank	0.000	07:17	
Control Test	0.080	07:17	
Air Blank	0.000	07:18	
Diagnostics Check	OK	07:18	

Data Download Date/Time: 07/28/2019 10:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,SCOTT,J
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: LAVERNE,DANIEL
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: TROTTER

DATE/TIME OF TEST: 07/12/2019 19:42:05

Diagnostics Check	OK	19:35
Air Blank	0.000	19:36
Control Test	0.080	19:36
Air Blank	0.000	19:36
Subject Sample # 1	0.234	19:37
Breath Volume	1.671	19:37
Air Blank	0.000	19:37
Air Blank	0.000	19:39
Subject Sample # 2	0.225	19:40
Breath Volume	1.511	19:40
Air Blank	0.000	19:41
Control Test	0.080	19:41
Air Blank	0.000	19:42
Diagnostics Check	OK	19:42

Data Download Date/Time: 07/28/2019 10:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,SCOTT,J
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: COZZONE,GAETANO
ARRESTING OFFICER AGENCY: Flagler Beach PD
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: ROJAS

DATE/TIME OF TEST: 07/13/2019 03:01:47

Diagnostics Check	OK	02:55
Air Blank	0.000	02:56
Control Test	0.080	02:56
Air Blank	0.000	02:56
Subject Sample # 1	0.130	02:57
Breath Volume	2.640	02:57
Air Blank	0.000	02:57
Air Blank	0.000	02:59
Subject Sample # 2	0.128	03:00
Breath Volume	2.175	03:00
Air Blank	0.000	03:00
Control Test	0.080	03:01
Air Blank	0.000	03:01
Diagnostics Check	OK	03:01

Data Download Date/Time: 07/28/2019 10:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,SCOTT,J
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: MCCARTHY,EVAN
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: BERTRAM

DATE/TIME OF TEST: 07/13/2019 05:42:14

Diagnostics Check	OK	05:35
Air Blank	0.000	05:35
Control Test	0.080	05:35
Air Blank	0.000	05:36
Subject Sample # 1	0.160	05:37
Breath Volume	1.515	05:37
Air Blank	0.000	05:37
Air Blank	0.000	05:39
Subject Sample # 2	0.155	05:40
Breath Volume	1.207	05:40
Air Blank	0.000	05:41
Control Test	0.079	05:41
Air Blank	0.000	05:42
Diagnostics Check	OK	05:42

Data Download Date/Time: 07/28/2019 10:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEEHAN,SHANE,P
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: MEEHAN,S
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: SARGENT

DATE/TIME OF TEST: 07/14/2019 01:20:42

Diagnostics Check	OK	01:12
Air Blank	0.000	01:12
Control Test	0.081	01:13
Air Blank	0.000	01:13
Subject Sample # 1	0.132	01:16
Breath Volume	1.531	01:16
Air Blank	0.000	01:16
Air Blank	0.000	01:18
Subject Sample # 2	0.146	01:19
Breath Volume	1.671	01:19
Air Blank	0.000	01:19
Control Test	0.080	01:20
Air Blank	0.000	01:20
Diagnostics Check	OK	01:20

Data Download Date/Time: 07/28/2019 10:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,SCOTT,J
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: LICHTY,DAVID
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 12:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: LUCERO

DATE/TIME OF TEST: 07/15/2019 01:45:23

Diagnostics Check	OK	01:34	
Air Blank	0.000	01:34	
Control Test	0.080	01:34	
Air Blank	0.000	01:35	
Subject Sample	# 1	01:38	No Sample Provided
Breath Volume	0.000	01:38	
Air Blank	0.000	01:39	
Air Blank	0.000	01:40	
Subject Sample	# 2	01:44	No Sample Provided
Breath Volume	0.000	01:44	
Air Blank	0.000	01:44	
Control Test	0.081	01:44	
Air Blank	0.000	01:45	
Diagnostics Check	OK	01:45	

Data Download Date/Time: 07/28/2019 10:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCARTHY, EVAN, V
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: GADDIE, KYLE
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: SHIMMEL

DATE/TIME OF TEST: 07/15/2019 03:42:47

Diagnostics Check	OK	03:35
Air Blank	0.000	03:36
Control Test	0.080	03:36
Air Blank	0.000	03:36
Subject Sample # 1	0.334	03:37
Breath Volume	3.277	03:37
Air Blank	0.000	03:38
Air Blank	0.000	03:40
Subject Sample # 2	0.327	03:41
Breath Volume	3.207	03:41
Air Blank	0.000	03:41
Control Test	0.079	03:42
Air Blank	0.000	03:42
Diagnostics Check	OK	03:42

Data Download Date/Time: 07/28/2019 10:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROUCH,JOSHUA,L
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: GOSSETT,A
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:46

SUBJECT'S LAST NAME: LAFOREST

DATE/TIME OF TEST: 07/15/2019 19:18:52

Diagnostics Check	OK	19:06	
Air Blank	0.000	19:07	
Control Test	0.081	19:07	
Air Blank	0.000	19:07	
Subject Sample	# 1 0.185	19:09	
Breath Volume	1.683	19:09	
Air Blank	0.000	19:09	
Air Blank	0.000	19:11	
Subject Sample	# 2 0.164	19:14	
Breath Volume	1.531	19:14	
Air Blank	0.000	19:14	
Air Blank	0.000	19:16	
Subject Sample	# 3	19:17	Subject Test Refused
Breath Volume	0.183	19:17	
Air Blank	0.000	19:17	
Control Test	0.079	19:18	
Air Blank	0.000	19:18	
Diagnostics Check	OK	19:18	

Data Download Date/Time: 07/28/2019 10:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,JASON,G
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: BROWN,JASON
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/16/2019 17:19:09

Diagnostics Check	OK	17:12
Air Blank	0.000	17:13
Control Test	0.081	17:13
Air Blank	0.000	17:14
Subject Sample # 1	0.000	17:14
Breath Volume	2.648	17:14
Air Blank	0.000	17:15
Air Blank	0.000	17:17
Subject Sample # 2	0.000	17:17
Breath Volume	3.062	17:17
Air Blank	0.000	17:18
Control Test	0.081	17:18
Air Blank	0.000	17:19
Diagnostics Check	OK	17:19

Data Download Date/Time: 07/28/2019 10:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,SCOTT,J
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: COZZONE,GAETANO
ARRESTING OFFICER AGENCY: Flagler Beach PD
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: MCMANUS

DATE/TIME OF TEST: 07/24/2019 02:51:44

Diagnostics Check	OK	02:40	
Air Blank	0.000	02:40	
Control Test	0.081	02:41	
Air Blank	0.000	02:41	
Subject Sample	# 1	02:44	No Sample Provided
Breath Volume	0.000	02:44	
Air Blank	0.000	02:45	
Air Blank	0.000	02:47	
Subject Sample	# 2	02:50	No Sample Provided
Breath Volume	0.000	02:50	
Air Blank	0.000	02:50	
Control Test	0.080	02:51	
Air Blank	0.000	02:51	
Diagnostics Check	OK	02:51	

Data Download Date/Time: 07/28/2019 10:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,SCOTT,J
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: DENKER,BRYAN
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: MUSIKAR

DATE/TIME OF TEST: 07/24/2019 03:47:09

Diagnostics Check	OK	03:40
Air Blank	0.000	03:41
Control Test	0.080	03:41
Air Blank	0.000	03:42
Subject Sample # 1	0.095	03:42
Breath Volume	3.468	03:42
Air Blank	0.000	03:43
Air Blank	0.000	03:45
Subject Sample # 2	0.095	03:45
Breath Volume	3.738	03:45
Air Blank	0.000	03:46
Control Test	0.080	03:46
Air Blank	0.000	03:47
Diagnostics Check	OK	03:47

Data Download Date/Time: 07/28/2019 10:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NICOLAS, MITA,
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: SHAUGHNESSY, GARRED
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: MORIARTY

DATE/TIME OF TEST: 07/26/2019 00:38:41

Diagnostics Check	OK	00:35	
Air Blank	0.000	00:36	
Control Test	0.081	00:36	
Air Blank	0.000	00:37	
Subject Sample	# 1	00:37	Subject Test Refused
Breath Volume	0.000	00:37	
Air Blank	0.000	00:37	
Control Test	0.080	00:38	
Air Blank	0.000	00:38	
Diagnostics Check	OK	00:38	

Data Download Date/Time: 07/28/2019 10:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCARTHY, EVAN, V
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: GADDIE, KYLE
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: SIMS

DATE/TIME OF TEST: 07/27/2019 03:16:11

Diagnostics Check		OK	03:06
Air Blank		0.000	03:06
Control Test		0.081	03:06
Air Blank		0.000	03:07
Subject Sample	# 1	0.258	03:08
Breath Volume		2.910	03:08
Air Blank		0.000	03:08
Air Blank		0.000	03:10
Subject Sample	# 2	0.290	03:11
Breath Volume		3.101	03:11
Air Blank		0.000	03:12
Air Blank		0.000	03:14
Subject Sample	# 3	0.280	03:14
Breath Volume		2.839	03:14
Air Blank		0.000	03:15
Control Test		0.080	03:15
Air Blank		0.000	03:16
Diagnostics Check		OK	03:16

Data Download Date/Time: 07/28/2019 10:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,CARL,L
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: PALMA,ANTONIO
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 06:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:58

SUBJECT'S LAST NAME: FLORENCE

DATE/TIME OF TEST: 07/27/2019 08:25:32

Diagnostics Check	OK	08:19
Air Blank	0.000	08:19
Control Test	0.081	08:19
Air Blank	0.000	08:20
Subject Sample # 1	0.081	08:21
Breath Volume	1.917	08:21
Air Blank	0.000	08:21
Air Blank	0.000	08:23
Subject Sample # 2	0.083	08:24
Breath Volume	1.433	08:24
Air Blank	0.000	08:24
Control Test	0.080	08:24
Air Blank	0.000	08:25
Diagnostics Check	OK	08:25

Data Download Date/Time: 07/28/2019 10:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: MANATEE COUNTY SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: MANATEE COUNTY SO
LAST AGENCY INSPECTION: 07/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: ROSSON

DATE/TIME OF TEST: 07/29/2019 00:01:45

Diagnostics Check	OK	23:55
Air Blank	0.000	23:55
Control Test	0.079	23:56
Air Blank	0.000	23:56
Subject Sample # 2	0.262	23:56
Breath Volume	1.968	23:56
Air Blank	0.000	23:57
Air Blank	0.000	23:59
Subject Sample # 1	0.242	00:00
Breath Volume	1.394	00:00
Air Blank	0.000	00:00
Control Test	0.078	00:01
Air Blank	0.000	00:01
Diagnostics Check	OK	00:01

Data Download Date/Time: 08/28/2019 09:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: MANATEE COUNTY SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: MANATEE COUNTY SO
LAST AGENCY INSPECTION: 07/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 19:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:14

SUBJECT'S LAST NAME: AFFOLTER

DATE/TIME OF TEST: 07/30/2019 20:41:01

Diagnostics Check	OK	20:34
Air Blank	0.000	20:35
Control Test	0.080	20:35
Air Blank	0.000	20:36
Subject Sample # 1	0.000	20:36
Breath Volume	3.382	20:36
Air Blank	0.000	20:37
Air Blank	0.000	20:39
Subject Sample # 2	0.000	20:39
Breath Volume	2.523	20:39
Air Blank	0.000	20:40
Control Test	0.078	20:40
Air Blank	0.000	20:40
Diagnostics Check	OK	20:41

Data Download Date/Time: 08/28/2019 09:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KALLMEYER,DANIEL,K
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: N/A,N/A
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 06/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: PROB/PAROL
ARREST TIME: 00:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:30

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 07/02/2019 13:18:06

Diagnostics Check	OK	13:12
Air Blank	0.000	13:12
Control Test	0.083	13:12
Air Blank	0.000	13:13
Subject Sample # 1	0.000	13:13
Breath Volume	2.144	13:13
Air Blank	0.000	13:14
Air Blank	0.000	13:16
Subject Sample # 2	0.000	13:16
Breath Volume	2.121	13:16
Air Blank	0.000	13:17
Control Test	0.083	13:17
Air Blank	0.000	13:18
Diagnostics Check	OK	13:18

Data Download Date/Time: 07/25/2019 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCKER, WILLIAM,
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: DELIMA, EDUARDO
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 06/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 07/05/2019 02:38:25

Diagnostics Check	OK	02:32
Air Blank	0.000	02:32
Control Test	0.083	02:32
Air Blank	0.000	02:33
Subject Sample # 1	0.186	02:33
Breath Volume	3.023	02:33
Air Blank	0.000	02:34
Air Blank	0.000	02:36
Subject Sample # 2	0.181	02:36
Breath Volume	3.195	02:36
Air Blank	0.000	02:37
Control Test	0.082	02:37
Air Blank	0.000	02:38
Diagnostics Check	OK	02:38

Data Download Date/Time: 07/25/2019 15:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KALLMEYER,DANIEL,K
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: N/A,N/A
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 06/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: PROB/PAROL
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:30

SUBJECT'S LAST NAME: ROMAN ROSA

DATE/TIME OF TEST: 07/09/2019 10:42:24

Diagnostics Check	OK	10:36
Air Blank	0.000	10:36
Control Test	0.083	10:37
Air Blank	0.000	10:37
Subject Sample # 1	0.000	10:38
Breath Volume	2.574	10:38
Air Blank	0.000	10:38
Air Blank	0.000	10:40
Subject Sample # 2	0.000	10:41
Breath Volume	2.746	10:41
Air Blank	0.000	10:41
Control Test	0.082	10:41
Air Blank	0.000	10:42
Diagnostics Check	OK	10:42

Data Download Date/Time: 07/25/2019 15:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KALLMEYER,DANIEL,K
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: N/A,N/A
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 06/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: PROB/PAROL
ARREST TIME: 00:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:45

SUBJECT'S LAST NAME: MORALES GARCIA

DATE/TIME OF TEST: 07/09/2019 13:21:28

Diagnostics Check	OK	13:15
Air Blank	0.000	13:15
Control Test	0.082	13:16
Air Blank	0.000	13:16
Subject Sample # 1	0.000	13:17
Breath Volume	1.929	13:17
Air Blank	0.000	13:17
Air Blank	0.000	13:19
Subject Sample # 2	0.000	13:20
Breath Volume	1.597	13:20
Air Blank	0.000	13:20
Control Test	0.082	13:20
Air Blank	0.000	13:21
Diagnostics Check	OK	13:21

Data Download Date/Time: 07/25/2019 15:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: BRUNER,SEAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 06/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:42

SUBJECT'S LAST NAME: BOTET

DATE/TIME OF TEST: 07/21/2019 21:10:10

Diagnostics Check	OK	21:03
Air Blank	0.000	21:04
Control Test	0.082	21:04
Air Blank	0.000	21:05
Subject Sample # 1	0.000	21:05
Breath Volume	3.496	21:05
Air Blank	0.000	21:06
Air Blank	0.000	21:08
Subject Sample # 2	0.000	21:08
Breath Volume	3.300	21:08
Air Blank	0.000	21:09
Control Test	0.084	21:09
Air Blank	0.000	21:10
Diagnostics Check	OK	21:10

Data Download Date/Time: 07/25/2019 15:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001152

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAGE,BRIAN,R
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: RIERA,SCOTT
ARRESTING OFFICER AGENCY: New Smyrna Beach PD
LAST AGENCY INSPECTION: 06/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A5
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: NORTON

DATE/TIME OF TEST: 07/04/2019 23:51:25

Diagnostics Check	OK	23:48	
Air Blank	0.000	23:49	
Control Test	0.081	23:49	
Air Blank	0.000	23:49	
Subject Sample	# 1	23:50	Subject Test Refused
Breath Volume	0.000	23:50	
Air Blank	0.000	23:50	
Control Test	0.080	23:50	
Air Blank	0.000	23:51	
Diagnostics Check	OK	23:51	

Data Download Date/Time: 07/18/2019 18:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001152

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAGE,BRIAN,R
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: RIERA,SCOTT
ARRESTING OFFICER AGENCY: New Smyrna Beach PD
LAST AGENCY INSPECTION: 06/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A5
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: WHITTIEMORE

DATE/TIME OF TEST: 07/13/2019 00:24:17

Diagnostics Check		OK	00:18
Air Blank		0.000	00:18
Control Test		0.079	00:18
Air Blank		0.000	00:19
Subject Sample	# 1	0.288	00:19
Breath Volume		1.339	00:19
Air Blank		0.000	00:20
Air Blank		0.000	00:22
Subject Sample	# 2	0.292	00:22
Breath Volume		1.945	00:22
Air Blank		0.000	00:23
Control Test		0.080	00:23
Air Blank		0.000	00:24
Diagnostics Check		OK	00:24

Data Download Date/Time: 07/18/2019 18:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001152

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAGE,BRIAN,R
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: JARRELL,JASSAMINE
ARRESTING OFFICER AGENCY: New Smyrna Beach PD
LAST AGENCY INSPECTION: 06/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A5
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: SOPHA

DATE/TIME OF TEST: 07/14/2019 02:19:08

Diagnostics Check	OK	02:09
Air Blank	0.000	02:10
Control Test	0.080	02:10
Air Blank	0.000	02:11
Subject Sample # 1	0.167	02:11
Breath Volume	1.144	02:11
Air Blank	0.000	02:12
Air Blank	0.000	02:14
Subject Sample # 2	0.189	02:14
Breath Volume	1.824	02:14
Air Blank	0.000	02:15
Air Blank	0.000	02:17
Subject Sample # 3	0.172	02:17
Breath Volume	1.500	02:17
Air Blank	0.000	02:18
Control Test	0.079	02:18
Air Blank	0.000	02:19
Diagnostics Check	OK	02:19

Data Download Date/Time: 07/18/2019 18:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001152

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAGE,BRIAN,R
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: CHRISTIE,CARLEY
ARRESTING OFFICER AGENCY: OTHER Fed
LAST AGENCY INSPECTION: 07/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A5
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:52

SUBJECT'S LAST NAME: CLAYTON

DATE/TIME OF TEST: 07/28/2019 20:19:36

Diagnostics Check	OK	20:12
Air Blank	0.000	20:13
Control Test	0.080	20:13
Air Blank	0.000	20:14
Subject Sample # 1	0.170	20:15
Breath Volume	3.269	20:15
Air Blank	0.000	20:15
Air Blank	0.000	20:17
Subject Sample # 2	0.168	20:18
Breath Volume	3.187	20:18
Air Blank	0.000	20:18
Control Test	0.079	20:18
Air Blank	0.000	20:19
Diagnostics Check	OK	20:19

Data Download Date/Time: 08/25/2019 08:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATSCH,JOSHUA,K
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: PATSCH,JOSHUA
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 07/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 12:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:23

SUBJECT'S LAST NAME: MURDOCK

DATE/TIME OF TEST: 07/14/2019 12:54:14

Diagnostics Check	OK	12:45	
Air Blank	0.000	12:45	
Control Test	0.082	12:46	
Air Blank	0.000	12:46	
Subject Sample # 1	0.000	12:49	
Breath Volume	1.406	12:49	
Air Blank	0.000	12:49	
Air Blank	0.000	12:51	
Subject Sample # 2	0.000	12:53	
Breath Volume	1.566	12:53	
Air Blank		12:53	Improper Sample
Air Blank	0.000	12:54	

Data Download Date/Time: 09/03/2019 09:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATSCH,JOSHUA,K
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: POWERS,STEVEN
ARRESTING OFFICER AGENCY: OTHER Fed
LAST AGENCY INSPECTION: 07/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 16:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:16

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 07/14/2019 17:43:47

Diagnostics Check	OK	17:37
Air Blank	0.000	17:38
Control Test	0.082	17:38
Air Blank	0.000	17:38
Subject Sample # 1	0.059	17:39
Breath Volume	3.957	17:39
Air Blank	0.000	17:40
Air Blank	0.000	17:41
Subject Sample # 2	0.056	17:42
Breath Volume	2.742	17:42
Air Blank	0.000	17:42
Control Test	0.081	17:43
Air Blank	0.000	17:43
Diagnostics Check	OK	17:43

Data Download Date/Time: 09/03/2019 09:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001155

INSTRUMENT REGISTERED TO: EGLIN AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOWELL, JONATHAN, A
BREATH TEST OPERATOR AGENCY: Eglin AFB PD
ARRESTING OFFICER: VIOLETTE, TYLER
ARRESTING OFFICER AGENCY: Eglin AFB PD
LAST AGENCY INSPECTION: 06/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:39

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 07/07/2019 22:26:33

Diagnostics Check	OK	22:25	
Air Blank		22:26	RFI Detect
Air Blank	0.000	22:26	

Data Download Date/Time: 07/08/2019 14:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001155

INSTRUMENT REGISTERED TO: EGLIN AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOWELL, JONATHAN, A
BREATH TEST OPERATOR AGENCY: Eglin AFB PD
ARRESTING OFFICER: VIOLETTE, TYLER
ARRESTING OFFICER AGENCY: Eglin AFB PD
LAST AGENCY INSPECTION: 06/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:39

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 07/07/2019 22:34:49

Diagnostics Check	OK	22:30	
Air Blank	0.000	22:31	
Control Test	0.079	22:31	
Air Blank	0.000	22:32	
Subject Sample	# 1	22:34	Sequence Aborted
Breath Volume	0.000	22:34	
Air Blank	0.000	22:34	

Data Download Date/Time: 07/08/2019 14:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001155

INSTRUMENT REGISTERED TO: EGLIN AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOWELL, JONATHAN, A
BREATH TEST OPERATOR AGENCY: Eglin AFB PD
ARRESTING OFFICER: VIOLETTE, TYLER
ARRESTING OFFICER AGENCY: Eglin AFB PD
LAST AGENCY INSPECTION: 06/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:39

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 07/07/2019 22:47:37

Diagnostics Check	OK	22:40
Air Blank	0.000	22:40
Control Test	0.078	22:41
Air Blank	0.000	22:41
Subject Sample # 1	0.311	22:43
Breath Volume	3.179	22:43
Air Blank	0.000	22:43
Air Blank	0.000	22:45
Subject Sample # 2	0.309	22:46
Breath Volume	1.847	22:46
Air Blank	0.000	22:46
Control Test	0.078	22:47
Air Blank	0.000	22:47
Diagnostics Check	OK	22:47

Data Download Date/Time: 07/08/2019 14:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001155

INSTRUMENT REGISTERED TO: EGLIN AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASEY, JOSHUA, P
BREATH TEST OPERATOR AGENCY: Eglin AFB PD
ARRESTING OFFICER: TOMASZEWSKI, TIMOTHY
ARRESTING OFFICER AGENCY: Eglin AFB PD
LAST AGENCY INSPECTION: 07/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: OSHEA

DATE/TIME OF TEST: 07/19/2019 02:02:56

Diagnostics Check	OK	01:56
Air Blank	0.000	01:56
Control Test	0.079	01:57
Air Blank	0.000	01:57
Subject Sample # 1	0.008	01:58
Breath Volume	1.386	01:58
Air Blank	0.000	01:59
Air Blank	0.000	02:01
Subject Sample # 2	0.000	02:01
Breath Volume	1.476	02:01
Air Blank	0.000	02:02
Control Test	0.078	02:02
Air Blank	0.000	02:02
Diagnostics Check	OK	02:02

Data Download Date/Time: 08/12/2019 13:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001155

INSTRUMENT REGISTERED TO: EGLIN AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASEY, JOSHUA, P
BREATH TEST OPERATOR AGENCY: Eglin AFB PD
ARRESTING OFFICER: TEMPLETON, ANNA
ARRESTING OFFICER AGENCY: Eglin AFB PD
LAST AGENCY INSPECTION: 07/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:04

SUBJECT'S LAST NAME: SPRING

DATE/TIME OF TEST: 07/28/2019 20:05:35

Diagnostics Check		OK	19:58
Air Blank		0.000	19:59
Control Test		0.079	19:59
Air Blank		0.000	19:59
Subject Sample	# 1	0.208	20:00
Breath Volume		1.484	20:00
Air Blank		0.000	20:01
Air Blank		0.000	20:02
Subject Sample	# 2	0.207	20:03
Breath Volume		1.707	20:03
Air Blank		0.000	20:04
Control Test		0.079	20:04
Air Blank		0.000	20:05
Diagnostics Check		OK	20:05

Data Download Date/Time: 08/12/2019 13:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001155

INSTRUMENT REGISTERED TO: EGLIN AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLACKWELL,JAMES,M
BREATH TEST OPERATOR AGENCY: Eglin AFB PD
ARRESTING OFFICER: SMITH,JEREMY
ARRESTING OFFICER AGENCY: Eglin AFB PD
LAST AGENCY INSPECTION: 07/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: ADMIN
ARREST TIME: 14:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:05

SUBJECT'S LAST NAME: PAASCH

DATE/TIME OF TEST: 07/31/2019 15:33:37

Diagnostics Check	OK	15:27
Air Blank	0.000	15:27
Control Test	0.079	15:27
Air Blank	0.000	15:28
Subject Sample # 1	0.024	15:29
Breath Volume	1.527	15:29
Air Blank	0.000	15:29
Air Blank	0.000	15:31
Subject Sample # 2	0.024	15:32
Breath Volume	1.375	15:32
Air Blank	0.000	15:32
Control Test	0.078	15:33
Air Blank	0.000	15:33
Diagnostics Check	OK	15:33

Data Download Date/Time: 08/12/2019 13:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,RICHARD,L
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: LONG,AUSTIN
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: LEYLAND

DATE/TIME OF TEST: 07/01/2019 01:23:25

Diagnostics Check		OK	01:16
Air Blank		0.000	01:17
Control Test		0.080	01:17
Air Blank		0.000	01:18
Subject Sample	# 1	0.207	01:18
Breath Volume		2.027	01:18
Air Blank		0.000	01:19
Air Blank		0.000	01:21
Subject Sample	# 2	0.212	01:21
Breath Volume		2.695	01:21
Air Blank		0.000	01:22
Control Test		0.081	01:22
Air Blank		0.000	01:23
Diagnostics Check		OK	01:23

Data Download Date/Time: 07/24/2019 09:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUBIN,RICHARD,D
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: RAIHALA,JON
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: PAGE

DATE/TIME OF TEST: 07/01/2019 23:35:27

Diagnostics Check	OK	23:32	
Air Blank	0.000	23:33	
Control Test	0.081	23:33	
Air Blank	0.000	23:33	
Subject Sample	# 1	23:34	Subject Test Refused
Breath Volume	0.000	23:34	
Air Blank	0.000	23:34	
Control Test	0.081	23:34	
Air Blank	0.000	23:35	
Diagnostics Check	OK	23:35	

Data Download Date/Time: 07/24/2019 09:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,RICHARD,L
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: GRISWOLD,ADAM
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: BUSING

DATE/TIME OF TEST: 07/05/2019 00:46:27

Diagnostics Check		OK	00:40
Air Blank		0.000	00:40
Control Test		0.081	00:41
Air Blank		0.000	00:41
Subject Sample	# 1	0.197	00:42
Breath Volume		3.457	00:42
Air Blank		0.000	00:42
Air Blank		0.000	00:44
Subject Sample	# 2	0.187	00:44
Breath Volume		3.222	00:44
Air Blank		0.000	00:45
Control Test		0.081	00:45
Air Blank		0.000	00:46
Diagnostics Check		OK	00:46

Data Download Date/Time: 07/24/2019 09:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARMINGTON, WILLIAM, G
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: RAIHALA, JON
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: HIGHTOWER

DATE/TIME OF TEST: 07/12/2019 02:25:12

Diagnostics Check	OK	02:22	
Air Blank	0.000	02:22	
Control Test	0.080	02:23	
Air Blank	0.000	02:23	
Subject Sample	# 1	02:23	Subject Test Refused
Breath Volume	0.000	02:23	
Air Blank	0.000	02:24	
Control Test	0.080	02:24	
Air Blank	0.000	02:25	
Diagnostics Check	OK	02:25	

Data Download Date/Time: 07/24/2019 09:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARNER, MICHAEL, D
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: LABRIE, DAVID
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 12:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:07

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 07/12/2019 13:37:52

Diagnostics Check	OK	13:34	
Air Blank	0.000	13:34	
Control Test	0.081	13:34	
Air Blank	0.000	13:35	
Subject Sample	# 1	13:36	Subject Test Refused
Breath Volume	0.000	13:36	
Air Blank	0.000	13:36	
Control Test	0.082	13:37	
Air Blank	0.000	13:37	
Diagnostics Check	OK	13:37	

Data Download Date/Time: 07/24/2019 09:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUBIN,RICHARD,D
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: HYATT,JUSTIN
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 07/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 17:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:42

SUBJECT'S LAST NAME: STILES

DATE/TIME OF TEST: 07/24/2019 19:09:26

Diagnostics Check	OK	19:02
Air Blank	0.000	19:03
Control Test	0.082	19:03
Air Blank	0.000	19:03
Subject Sample # 1	0.129	19:05
Breath Volume	2.324	19:05
Air Blank	0.000	19:05
Air Blank	0.000	19:07
Subject Sample # 2	0.128	19:07
Breath Volume	1.777	19:07
Air Blank	0.000	19:08
Control Test	0.081	19:08
Air Blank	0.000	19:09
Diagnostics Check	OK	19:09

Data Download Date/Time: 08/26/2019 12:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUBIN,RICHARD,D
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: RAIHALA,JON
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 07/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: TROTT

DATE/TIME OF TEST: 07/25/2019 02:26:38

Diagnostics Check	OK	02:20
Air Blank	0.000	02:20
Control Test	0.081	02:21
Air Blank	0.000	02:21
Subject Sample # 1	0.277	02:22
Breath Volume	1.851	02:22
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample # 2	0.269	02:25
Breath Volume	3.128	02:25
Air Blank	0.000	02:25
Control Test	0.081	02:26
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

Data Download Date/Time: 08/26/2019 12:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,RICHARD,L
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: LONG,AUSTIN
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 07/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: BURKE

DATE/TIME OF TEST: 07/26/2019 23:53:53

Diagnostics Check	OK	23:47
Air Blank	0.000	23:47
Control Test	0.081	23:47
Air Blank	0.000	23:48
Subject Sample # 1	0.000	23:49
Breath Volume	2.421	23:49
Air Blank	0.000	23:49
Air Blank	0.000	23:51
Subject Sample # 2	0.000	23:52
Breath Volume	3.093	23:52
Air Blank	0.000	23:52
Control Test	0.082	23:53
Air Blank	0.000	23:53
Diagnostics Check	OK	23:53

Data Download Date/Time: 08/26/2019 12:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001159

INSTRUMENT REGISTERED TO: ORANGE CITY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUBAKER,KELLY,J
BREATH TEST OPERATOR AGENCY: Orange City PD
ARRESTING OFFICER: THORNTON,KENNETH
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 07/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: BUI
ARREST TIME: 17:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: WALDOW

DATE/TIME OF TEST: 07/20/2019 19:43:58

Diagnostics Check	OK	19:37
Air Blank	0.000	19:37
Control Test	0.081	19:38
Air Blank	0.000	19:38
Subject Sample # 1	0.194	19:39
Breath Volume	2.867	19:39
Air Blank	0.000	19:40
Air Blank	0.000	19:41
Subject Sample # 2	0.192	19:42
Breath Volume	2.144	19:42
Air Blank	0.000	19:43
Control Test	0.081	19:43
Air Blank	0.000	19:43
Diagnostics Check	OK	19:43

Data Download Date/Time: 08/01/2019 14:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001190

INSTRUMENT REGISTERED TO: BOYNTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTRO,DENNIS,
BREATH TEST OPERATOR AGENCY: Boynton Beach PD
ARRESTING OFFICER: CASTRO,DENNIS
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: ALEXANDER

DATE/TIME OF TEST: 07/20/2019 01:08:51

Diagnostics Check	OK	01:06	
Air Blank	0.000	01:06	
Control Test	0.078	01:06	
Air Blank	0.000	01:07	
Subject Sample	# 1	01:07	Subject Test Refused
Breath Volume	0.000	01:07	
Air Blank	0.000	01:07	
Control Test	0.077	01:08	
Air Blank	0.000	01:08	
Diagnostics Check	OK	01:08	

Data Download Date/Time: 08/03/2019 22:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001190

INSTRUMENT REGISTERED TO: BOYNTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTRO,DENNIS,
BREATH TEST OPERATOR AGENCY: Boynton Beach PD
ARRESTING OFFICER: CASTRO,DENNIS
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 15:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:30

SUBJECT'S LAST NAME: LEHMAN

DATE/TIME OF TEST: 07/24/2019 16:02:13

Diagnostics Check	OK	15:59	
Air Blank	0.000	15:59	
Control Test	0.078	16:00	
Air Blank	0.000	16:00	
Subject Sample	# 1	16:00	Subject Test Refused
Breath Volume	0.000	16:00	
Air Blank	0.000	16:01	
Control Test	0.077	16:01	
Air Blank	0.000	16:02	
Diagnostics Check	OK	16:02	

Data Download Date/Time: 08/03/2019 22:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TARR,L,J
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: BOGART,C
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 07/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: CONTRERAS

DATE/TIME OF TEST: 07/02/2019 03:52:55

Diagnostics Check	OK	03:44
Air Blank	0.000	03:45
Control Test	0.079	03:45
Air Blank	0.000	03:46
Subject Sample # 1	0.215	03:47
Breath Volume	1.539	03:47
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 2	0.223	03:51
Breath Volume	2.460	03:51
Air Blank	0.000	03:52
Control Test	0.078	03:52
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time: 08/01/2019 12:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,L
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: KENNEDY,DANIEL
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 07/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: JAVORSKY

DATE/TIME OF TEST: 07/03/2019 02:34:34

Diagnostics Check	OK	02:29	
Air Blank	0.000	02:30	
Control Test	0.078	02:30	
Air Blank	0.000	02:31	
Subject Sample	# 1	02:33	Subject Test Refused
Breath Volume	0.000	02:33	
Air Blank	0.000	02:33	
Control Test	0.079	02:33	
Air Blank	0.000	02:34	
Diagnostics Check	OK	02:34	

Data Download Date/Time: 08/01/2019 12:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONSEES,D,T
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: POUNDS,R
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 07/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: RZASA

DATE/TIME OF TEST: 07/04/2019 22:30:16

Diagnostics Check	OK	22:23
Air Blank	0.000	22:23
Control Test	0.078	22:24
Air Blank	0.000	22:24
Subject Sample # 1	0.252	22:25
Breath Volume	1.160	22:25
Air Blank	0.000	22:25
Air Blank	0.000	22:27
Subject Sample # 2	0.263	22:28
Breath Volume	2.226	22:28
Air Blank	0.000	22:29
Control Test	0.078	22:29
Air Blank	0.000	22:30
Diagnostics Check	OK	22:30

Data Download Date/Time: 08/01/2019 12:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOSEPH,CUTHBERT,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: SCHCICHTE,J
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 07/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: ALVARADO BODON

DATE/TIME OF TEST: 07/06/2019 20:12:19

Diagnostics Check	OK	20:05
Air Blank	0.000	20:06
Control Test	0.079	20:06
Air Blank	0.000	20:06
Subject Sample # 1	0.145	20:07
Breath Volume	1.796	20:07
Air Blank	0.000	20:08
Air Blank	0.000	20:10
Subject Sample # 2	0.149	20:10
Breath Volume	1.511	20:10
Air Blank	0.000	20:11
Control Test	0.078	20:11
Air Blank	0.000	20:12
Diagnostics Check	OK	20:12

Data Download Date/Time: 08/01/2019 12:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,S,C
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: NOVA,M
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 07/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: HARPER

DATE/TIME OF TEST: 07/07/2019 01:24:54

Diagnostics Check	OK	01:18
Air Blank	0.000	01:19
Control Test	0.079	01:19
Air Blank	0.000	01:19
Subject Sample # 1	0.268	01:20
Breath Volume	2.164	01:20
Air Blank	0.000	01:21
Air Blank	0.000	01:22
Subject Sample # 2	0.272	01:23
Breath Volume	1.957	01:23
Air Blank	0.000	01:24
Control Test	0.078	01:24
Air Blank	0.000	01:24
Diagnostics Check	OK	01:24

Data Download Date/Time: 08/01/2019 12:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUGO,JONATHAN,B
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: MOLCAN,M
ARRESTING OFFICER AGENCY: NAPLES PD
LAST AGENCY INSPECTION: 07/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:26

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 07/07/2019 22:07:05

Diagnostics Check	OK	21:55	
Air Blank	0.000	21:55	
Control Test	0.078	21:56	
Air Blank	0.000	21:56	
Subject Sample	# 1	22:00	No Sample Provided
Breath Volume	0.000	22:00	
Air Blank	0.000	22:00	
Air Blank	0.000	22:02	
Subject Sample	# 2	22:05	No Sample Provided
Breath Volume	0.000	22:05	
Air Blank	0.000	22:06	
Control Test	0.079	22:06	
Air Blank	0.000	22:06	
Diagnostics Check	OK	22:07	

Data Download Date/Time: 08/01/2019 12:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEORGE,SHANNON,N
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: ROWLEY,C
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 07/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: SICKLER

DATE/TIME OF TEST: 07/11/2019 04:17:58

Diagnostics Check	OK	04:09
Air Blank	0.000	04:10
Control Test	0.078	04:10
Air Blank	0.000	04:10
Subject Sample # 1	0.131	04:12
Breath Volume	1.515	04:12
Air Blank	0.000	04:12
Air Blank	0.000	04:14
Subject Sample # 2	0.126	04:16
Breath Volume	2.093	04:16
Air Blank	0.000	04:17
Control Test	0.079	04:17
Air Blank	0.000	04:17
Diagnostics Check	OK	04:17

Data Download Date/Time: 08/01/2019 12:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODI,DOMENIC,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: HENDERSON,K
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 07/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:28

SUBJECT'S LAST NAME: CORREALE

DATE/TIME OF TEST: 07/11/2019 19:06:55

Diagnostics Check	OK	18:55	
Air Blank	0.000	18:55	
Control Test	0.078	18:56	
Air Blank	0.000	18:56	
Subject Sample	# 1 0.233	18:59	Volume Not Met
Breath Volume	0.820	18:59	
Air Blank	0.000	19:00	
Air Blank	0.000	19:02	
Subject Sample	# 2 0.213	19:05	Slope Not Level
Breath Volume	1.203	19:05	
Air Blank	0.000	19:06	
Control Test	0.077	19:06	
Air Blank	0.000	19:06	
Diagnostics Check	OK	19:06	

Data Download Date/Time: 08/01/2019 12:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODI,DOMENIC,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: HENDERSON,K
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 07/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:28

SUBJECT'S LAST NAME: CORREALE

DATE/TIME OF TEST: 07/11/2019 19:20:18

Diagnostics Check	OK	19:09	
Air Blank	0.000	19:10	
Control Test	0.078	19:10	
Air Blank	0.000	19:10	
Subject Sample	# 1 0.231	19:13	
Breath Volume	1.125	19:13	
Air Blank	0.000	19:14	
Air Blank	0.000	19:15	
Subject Sample	# 2	19:18	Subject Test Refused
Breath Volume	1.191	19:18	
Air Blank	0.000	19:19	
Control Test	0.079	19:19	
Air Blank	0.000	19:20	
Diagnostics Check	OK	19:20	

Data Download Date/Time: 08/01/2019 12:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLANEDA ROMO,KARLA,A
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: GRANNMAN,RYAN
ARRESTING OFFICER AGENCY: MARCO ISLAND PD
LAST AGENCY INSPECTION: 07/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: MOJICA GASPER

DATE/TIME OF TEST: 07/13/2019 23:17:56

Diagnostics Check	OK	23:11
Air Blank	0.000	23:11
Control Test	0.078	23:12
Air Blank	0.000	23:12
Subject Sample # 1	0.276	23:13
Breath Volume	2.875	23:13
Air Blank	0.000	23:13
Air Blank	0.000	23:15
Subject Sample # 2	0.280	23:16
Breath Volume	2.476	23:16
Air Blank	0.000	23:17
Control Test	0.079	23:17
Air Blank	0.000	23:17
Diagnostics Check	OK	23:17

Data Download Date/Time: 08/01/2019 12:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUONPANE,V,D
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: MCCOY,J
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 07/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: BERNSEN

DATE/TIME OF TEST: 07/15/2019 03:00:07

Diagnostics Check	OK	02:53
Air Blank	0.000	02:53
Control Test	0.079	02:54
Air Blank	0.000	02:54
Subject Sample # 1	0.000	02:55
Breath Volume	2.289	02:55
Air Blank	0.000	02:56
Air Blank	0.000	02:58
Subject Sample # 2	0.000	02:58
Breath Volume	2.066	02:58
Air Blank	0.000	02:59
Control Test	0.079	02:59
Air Blank	0.000	03:00
Diagnostics Check	OK	03:00

Data Download Date/Time: 08/01/2019 12:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOPS,C,M
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: KAZMIERCZAK,B
ARRESTING OFFICER AGENCY: NAPLES PD
LAST AGENCY INSPECTION: 07/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: POMAKIS

DATE/TIME OF TEST: 07/16/2019 02:33:49

Diagnostics Check	OK	02:25
Air Blank	0.000	02:26
Control Test	0.079	02:26
Air Blank	0.000	02:27
Subject Sample # 1	0.196	02:28
Breath Volume	1.320	02:28
Air Blank	0.000	02:29
Air Blank	0.000	02:30
Subject Sample # 2	0.186	02:32
Breath Volume	1.437	02:32
Air Blank	0.000	02:32
Control Test	0.078	02:33
Air Blank	0.000	02:33
Diagnostics Check	OK	02:33

Data Download Date/Time: 08/01/2019 12:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORNELIUS,TIM,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: CORNELIUS,TIM
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 07/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: GONZALEZ MERLIN

DATE/TIME OF TEST: 07/19/2019 23:22:15

Diagnostics Check	OK	23:14
Air Blank	0.000	23:15
Control Test	0.078	23:15
Air Blank	0.000	23:16
Subject Sample # 1	0.197	23:17
Breath Volume	2.792	23:17
Air Blank	0.000	23:18
Air Blank	0.000	23:20
Subject Sample # 2	0.178	23:20
Breath Volume	1.539	23:20
Air Blank	0.000	23:21
Control Test	0.077	23:21
Air Blank	0.000	23:22
Diagnostics Check	OK	23:22

Data Download Date/Time: 08/01/2019 12:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOSEPH,CUTHBERT,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: BOGART,S.M
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 07/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: MORALES BAUTISTA

DATE/TIME OF TEST: 07/20/2019 00:54:21

Diagnostics Check	OK	00:47
Air Blank	0.000	00:48
Control Test	0.079	00:48
Air Blank	0.000	00:48
Subject Sample # 1	0.221	00:49
Breath Volume	2.445	00:49
Air Blank	0.000	00:50
Air Blank	0.000	00:52
Subject Sample # 2	0.224	00:52
Breath Volume	3.714	00:52
Air Blank	0.000	00:53
Control Test	0.078	00:53
Air Blank	0.000	00:54
Diagnostics Check	OK	00:54

Data Download Date/Time: 08/01/2019 12:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORNELIUS,TIM,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: CORNELIUS,TIM
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 07/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 07/20/2019 02:56:59

Diagnostics Check		OK	02:50
Air Blank		0.000	02:50
Control Test		0.079	02:51
Air Blank		0.000	02:51
Subject Sample	# 1	0.148	02:52
Breath Volume		2.019	02:52
Air Blank		0.000	02:53
Air Blank		0.000	02:54
Subject Sample	# 2	0.154	02:55
Breath Volume		2.347	02:55
Air Blank		0.000	02:56
Control Test		0.078	02:56
Air Blank		0.000	02:56
Diagnostics Check		OK	02:56

Data Download Date/Time: 08/01/2019 12:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NATSON,ANDRAE,M
BREATH TEST OPERATOR AGENCY: FHP TROOP F
ARRESTING OFFICER: SUEROLANTIGUA,J
ARRESTING OFFICER AGENCY: FHP TROOP F
LAST AGENCY INSPECTION: 07/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: BRITT

DATE/TIME OF TEST: 07/20/2019 21:58:02

Diagnostics Check	OK	21:52
Air Blank	0.000	21:52
Control Test	0.079	21:52
Air Blank	0.000	21:53
Subject Sample # 1	0.000	21:53
Breath Volume	1.835	21:53
Air Blank	0.000	21:54
Air Blank	0.000	21:56
Subject Sample # 2	0.000	21:56
Breath Volume	3.269	21:56
Air Blank	0.000	21:57
Control Test	0.079	21:57
Air Blank	0.000	21:57
Diagnostics Check	OK	21:58

Data Download Date/Time: 08/01/2019 12:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUGO,J,B
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: MENDOZA,U
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 07/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: BURGESS

DATE/TIME OF TEST: 07/22/2019 03:58:39

Diagnostics Check	OK	03:51
Air Blank	0.000	03:52
Control Test	0.078	03:52
Air Blank	0.000	03:53
Subject Sample # 1	0.067	03:53
Breath Volume	1.386	03:53
Air Blank	0.000	03:54
Air Blank	0.000	03:56
Subject Sample # 2	0.067	03:57
Breath Volume	1.398	03:57
Air Blank	0.000	03:57
Control Test	0.079	03:58
Air Blank	0.000	03:58
Diagnostics Check	OK	03:58

Data Download Date/Time: 08/01/2019 12:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,E,W
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: CORWIN,J
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 07/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: OLSEN

DATE/TIME OF TEST: 07/23/2019 02:31:15

Diagnostics Check	OK	02:23
Air Blank	0.000	02:24
Control Test	0.078	02:24
Air Blank	0.000	02:25
Subject Sample # 1	0.119	02:25
Breath Volume	1.402	02:25
Air Blank	0.000	02:26
Air Blank	0.000	02:28
Subject Sample # 2	0.122	02:29
Breath Volume	1.808	02:29
Air Blank	0.000	02:30
Control Test	0.079	02:30
Air Blank	0.000	02:31
Diagnostics Check	OK	02:31

Data Download Date/Time: 08/01/2019 12:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOPS,C,M
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: OLIVASTRO,N
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 07/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: JACOBS

DATE/TIME OF TEST: 07/24/2019 21:10:55

Diagnostics Check	OK	21:04
Air Blank	0.000	21:05
Control Test	0.079	21:05
Air Blank	0.000	21:06
Subject Sample # 1	0.137	21:06
Breath Volume	2.332	21:06
Air Blank	0.000	21:07
Air Blank	0.000	21:08
Subject Sample # 2	0.141	21:09
Breath Volume	2.113	21:09
Air Blank	0.000	21:10
Control Test	0.079	21:10
Air Blank	0.000	21:10
Diagnostics Check	OK	21:10

Data Download Date/Time: 08/01/2019 12:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,ZACHARY,H
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: SCHAARE,D
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 07/01/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 14:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: STANTON

DATE/TIME OF TEST: 07/25/2019 16:56:55

Diagnostics Check	OK	16:50
Air Blank	0.000	16:51
Control Test	0.078	16:51
Air Blank	0.000	16:52
Subject Sample # 1	0.244	16:52
Breath Volume	1.484	16:52
Air Blank	0.000	16:53
Air Blank	0.000	16:54
Subject Sample # 2	0.233	16:55
Breath Volume	2.210	16:55
Air Blank	0.000	16:56
Control Test	0.079	16:56
Air Blank	0.000	16:56
Diagnostics Check	OK	16:56

Data Download Date/Time: 08/01/2019 12:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS, MATTHEW, L
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: ALTER, A
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 07/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: MINTON

DATE/TIME OF TEST: 07/29/2019 04:29:00

Diagnostics Check	OK	04:22
Air Blank	0.000	04:23
Control Test	0.079	04:23
Air Blank	0.000	04:23
Subject Sample # 1	0.241	04:24
Breath Volume	3.324	04:24
Air Blank	0.000	04:25
Air Blank	0.000	04:26
Subject Sample # 2	0.240	04:27
Breath Volume	3.621	04:27
Air Blank	0.000	04:28
Control Test	0.078	04:28
Air Blank	0.000	04:28
Diagnostics Check	OK	04:29

Data Download Date/Time: 08/01/2019 12:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVAS,W,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: ALLEN,R
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 07/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 07/30/2019 04:13:59

Diagnostics Check	OK	04:07
Air Blank	0.000	04:07
Control Test	0.078	04:08
Air Blank	0.000	04:08
Subject Sample # 1	0.190	04:09
Breath Volume	3.054	04:09
Air Blank	0.000	04:10
Air Blank	0.000	04:11
Subject Sample # 2	0.190	04:12
Breath Volume	2.843	04:12
Air Blank	0.000	04:13
Control Test	0.078	04:13
Air Blank	0.000	04:13
Diagnostics Check	OK	04:13

Data Download Date/Time: 08/01/2019 12:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANNORAT,ADELY,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: WOODS,H
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 07/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: BACIK

DATE/TIME OF TEST: 07/31/2019 21:31:26

Diagnostics Check	OK	21:25
Air Blank	0.000	21:25
Control Test	0.079	21:26
Air Blank	0.000	21:26
Subject Sample # 1	0.114	21:26
Breath Volume	1.699	21:26
Air Blank	0.000	21:27
Air Blank	0.000	21:29
Subject Sample # 2	0.116	21:29
Breath Volume	2.160	21:29
Air Blank	0.000	21:30
Control Test	0.078	21:30
Air Blank	0.000	21:31
Diagnostics Check	OK	21:31

Data Download Date/Time: 08/01/2019 12:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,JONATHAN,R
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: TURNER,JONATHAN
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: FLORES

DATE/TIME OF TEST: 07/03/2019 00:39:07

Diagnostics Check	OK	00:32
Air Blank	0.000	00:33
Control Test	0.080	00:33
Air Blank	0.000	00:34
Subject Sample # 1	0.160	00:34
Breath Volume	1.457	00:34
Air Blank	0.000	00:35
Air Blank	0.000	00:37
Subject Sample # 2	0.165	00:37
Breath Volume	1.281	00:37
Air Blank	0.000	00:38
Control Test	0.080	00:38
Air Blank	0.000	00:39
Diagnostics Check	OK	00:39

Data Download Date/Time: 08/04/2019 19:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAGUADOG,SHAUN,J
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: JOHNSON,ROBERT
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 07/05/2019 01:17:29

Diagnostics Check		OK	01:10
Air Blank		0.000	01:11
Control Test		0.080	01:11
Air Blank		0.000	01:12
Subject Sample	# 1	0.218	01:13
Breath Volume		2.187	01:13
Air Blank		0.000	01:13
Air Blank		0.000	01:15
Subject Sample	# 2	0.212	01:16
Breath Volume		1.277	01:16
Air Blank		0.000	01:16
Control Test		0.079	01:16
Air Blank		0.000	01:17
Diagnostics Check		OK	01:17

Data Download Date/Time: 08/04/2019 19:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOUSE,STEVEN,D
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: HOUSE,STEVEN
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 04:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: SHORTT

DATE/TIME OF TEST: 07/14/2019 04:58:40

Diagnostics Check		OK	04:52
Air Blank		0.000	04:52
Control Test		0.080	04:53
Air Blank		0.000	04:53
Subject Sample	# 1	0.204	04:54
Breath Volume		4.964	04:54
Air Blank		0.000	04:54
Air Blank		0.000	04:56
Subject Sample	# 2	0.214	04:57
Breath Volume		3.632	04:57
Air Blank		0.000	04:57
Control Test		0.079	04:58
Air Blank		0.000	04:58
Diagnostics Check		OK	04:58

Data Download Date/Time: 08/04/2019 19:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAUGA,JOHN,E
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: HUTCHINSON,JOSEPH
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 05:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: BURKE

DATE/TIME OF TEST: 07/17/2019 05:41:56

Diagnostics Check	OK	05:36
Air Blank	0.000	05:36
Control Test	0.080	05:36
Air Blank	0.000	05:37
Subject Sample # 1	0.000	05:37
Breath Volume	2.902	05:37
Air Blank	0.000	05:38
Air Blank	0.000	05:40
Subject Sample # 2	0.000	05:40
Breath Volume	1.562	05:40
Air Blank	0.000	05:41
Control Test	0.080	05:41
Air Blank	0.000	05:41
Diagnostics Check	OK	05:41

Data Download Date/Time: 08/04/2019 19:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,JONATHAN,R
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: SMITH,CAITLIN
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:18

SUBJECT'S LAST NAME: SPINDLER

DATE/TIME OF TEST: 07/21/2019 21:51:06

Diagnostics Check		OK	21:38	
Air Blank		0.000	21:39	
Control Test		0.080	21:39	
Air Blank		0.000	21:40	
Subject Sample	# 1	0.142	21:43	Volume Not Met
Breath Volume		1.035	21:43	
Air Blank		0.000	21:43	
Air Blank		0.000	21:45	
Subject Sample	# 2	0.179	21:46	No .020 agreement
Breath Volume		1.132	21:46	
Air Blank		0.000	21:47	
Air Blank		0.000	21:48	
Subject Sample	# 3	0.153	21:49	No .020 agreement
Breath Volume		1.472	21:49	
Air Blank		0.000	21:50	
Control Test		0.079	21:50	
Air Blank		0.000	21:51	
Diagnostics Check		OK	21:51	

Data Download Date/Time: 08/04/2019 19:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMSEY,JOHNATHAN,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: MONAHAN,TIMOTHY
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 08:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:32

SUBJECT'S LAST NAME: BARKER

DATE/TIME OF TEST: 07/29/2019 09:00:23

Diagnostics Check	OK	08:53
Air Blank	0.000	08:53
Control Test	0.081	08:54
Air Blank	0.000	08:54
Subject Sample # 1	0.000	08:55
Breath Volume	2.933	08:55
Air Blank	0.000	08:56
Air Blank	0.000	08:58
Subject Sample # 2	0.000	08:59
Breath Volume	3.511	08:59
Air Blank	0.000	08:59
Control Test	0.080	08:59
Air Blank	0.000	09:00
Diagnostics Check	OK	09:00

Data Download Date/Time: 08/04/2019 19:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,JONATHAN,R
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: LANIER,T
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: LOWERY

DATE/TIME OF TEST: 07/30/2019 04:28:58

Diagnostics Check	OK	04:24	
Air Blank	0.000	04:25	
Control Test	0.080	04:25	
Air Blank	0.000	04:26	
Subject Sample	# 1	04:27	Subject Test Refused
Breath Volume	0.070	04:27	
Air Blank	0.000	04:28	
Control Test	0.080	04:28	
Air Blank	0.000	04:28	
Diagnostics Check	OK	04:28	

Data Download Date/Time: 08/04/2019 19:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001207

INSTRUMENT REGISTERED TO: HENDRY COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORWOOD,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: NORWOOD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 06/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: RUDD

DATE/TIME OF TEST: 07/05/2019 22:33:54

Diagnostics Check	OK	22:27
Air Blank	0.000	22:28
Control Test	0.081	22:28
Air Blank	0.000	22:28
Subject Sample # 1	0.122	22:29
Breath Volume	2.070	22:29
Air Blank	0.000	22:29
Air Blank	0.000	22:31
Subject Sample # 2	0.123	22:32
Breath Volume	2.218	22:32
Air Blank	0.000	22:32
Control Test	0.081	22:33
Air Blank	0.000	22:33
Diagnostics Check	OK	22:33

Data Download Date/Time: 07/12/2019 03:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001207

INSTRUMENT REGISTERED TO: HENDRY COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORWOOD,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: NORWOOD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 06/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: FRANCIS

DATE/TIME OF TEST: 07/07/2019 00:49:16

Diagnostics Check		OK	00:43
Air Blank		0.000	00:43
Control Test		0.082	00:43
Air Blank		0.000	00:44
Subject Sample	# 1	0.097	00:44
Breath Volume		2.003	00:44
Air Blank		0.000	00:45
Air Blank		0.000	00:47
Subject Sample	# 2	0.092	00:47
Breath Volume		2.074	00:47
Air Blank		0.000	00:48
Control Test		0.082	00:48
Air Blank		0.000	00:49
Diagnostics Check		OK	00:49

Data Download Date/Time: 07/12/2019 03:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001207

INSTRUMENT REGISTERED TO: HENDRY COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORWOOD,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: ANDERSON,KALEIGH
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 07/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: CHAIREZ

DATE/TIME OF TEST: 07/21/2019 01:40:28

Diagnostics Check	OK	01:34
Air Blank	0.000	01:34
Control Test	0.081	01:34
Air Blank	0.000	01:35
Subject Sample # 1	0.025	01:36
Breath Volume	1.253	01:36
Air Blank	0.000	01:36
Air Blank	0.000	01:38
Subject Sample # 2	0.025	01:38
Breath Volume	2.089	01:38
Air Blank	0.000	01:39
Control Test	0.081	01:39
Air Blank	0.000	01:40
Diagnostics Check	OK	01:40

Data Download Date/Time: 09/03/2019 12:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,LAUREN,A
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: SMITH,LAUREN
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 06/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: ADMIN
ARREST TIME: 11:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:53

SUBJECT'S LAST NAME: CAIN

DATE/TIME OF TEST: 07/03/2019 12:23:35

Diagnostics Check	OK	12:17
Air Blank	0.000	12:18
Control Test	0.080	12:18
Air Blank	0.000	12:18
Subject Sample # 1	0.116	12:19
Breath Volume	2.316	12:19
Air Blank	0.000	12:19
Air Blank	0.000	12:21
Subject Sample # 2	0.114	12:22
Breath Volume	2.726	12:22
Air Blank	0.000	12:22
Control Test	0.081	12:22
Air Blank	0.000	12:23
Diagnostics Check	OK	12:23

Data Download Date/Time: 07/29/2019 15:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,ELIZABETH,D
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: RODRIGUEZ,MARCELES
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 06/03/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: KUHN

DATE/TIME OF TEST: 07/04/2019 01:02:05

Diagnostics Check	OK	00:55
Air Blank	0.000	00:56
Control Test	0.078	00:56
Air Blank	0.000	00:57
Subject Sample # 1	0.178	00:57
Breath Volume	1.656	00:57
Air Blank	0.000	00:58
Air Blank	0.000	01:00
Subject Sample # 2	0.175	01:00
Breath Volume	1.371	01:00
Air Blank	0.000	01:01
Control Test	0.077	01:01
Air Blank	0.000	01:02
Diagnostics Check	OK	01:02

Data Download Date/Time: 07/29/2019 15:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,ELIZABETH,D
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: ADKINS,ANDREW
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 06/03/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: DREW

DATE/TIME OF TEST: 07/05/2019 02:41:50

Diagnostics Check	OK	02:35
Air Blank	0.000	02:35
Control Test	0.081	02:36
Air Blank	0.000	02:36
Subject Sample # 1	0.013	02:37
Breath Volume	4.058	02:37
Air Blank	0.000	02:37
Air Blank	0.000	02:39
Subject Sample # 2	0.012	02:40
Breath Volume	2.183	02:40
Air Blank	0.000	02:40
Control Test	0.081	02:41
Air Blank	0.000	02:41
Diagnostics Check	OK	02:41

Data Download Date/Time: 07/29/2019 15:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUBNIS,GUY,W
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: MCDOWELL,MICHAEL
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 06/03/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 03:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: PEEK

DATE/TIME OF TEST: 07/07/2019 04:29:34

Diagnostics Check	OK	04:22
Air Blank	0.000	04:23
Control Test	0.077	04:23
Air Blank	0.000	04:24
Subject Sample # 1	0.112	04:24
Breath Volume	1.453	04:24
Air Blank	0.000	04:25
Air Blank	0.000	04:27
Subject Sample # 2	0.116	04:28
Breath Volume	1.703	04:28
Air Blank	0.000	04:28
Control Test	0.077	04:28
Air Blank	0.000	04:29
Diagnostics Check	OK	04:29

Data Download Date/Time: 07/29/2019 15:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMBY,TAYLOR,N
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: SHEELY,DANTE
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 06/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: KUBAS

DATE/TIME OF TEST: 07/18/2019 23:23:01

Diagnostics Check	OK	23:14
Air Blank	0.000	23:15
Control Test	0.080	23:15
Air Blank	0.000	23:16
Subject Sample # 1	0.175	23:18
Breath Volume	1.937	23:18
Air Blank	0.000	23:19
Air Blank	0.000	23:20
Subject Sample # 2	0.161	23:21
Breath Volume	1.484	23:21
Air Blank	0.000	23:22
Control Test	0.080	23:22
Air Blank	0.000	23:22
Diagnostics Check	OK	23:23

Data Download Date/Time: 07/29/2019 15:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUBNIS,GUY,W
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: SIKES,ROBERT
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 06/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 07/29/2019 03:18:12

Diagnostics Check	OK	03:09	
Air Blank	0.000	03:10	
Control Test	0.081	03:10	
Air Blank	0.000	03:11	
Subject Sample # 1	0.169	03:13	
Breath Volume	1.304	03:13	
Air Blank	0.000	03:14	
Air Blank	0.000	03:16	
Subject Sample # 2	0.158	03:16	
Breath Volume	1.304	03:16	
Air Blank	0.000	03:17	
Control Test	0.073	03:17	Control Outside Tolerance
Air Blank	0.000	03:18	

Data Download Date/Time: 07/29/2019 15:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUBNIS,GUY,W
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: SIKES,ROBERT
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 06/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 07/29/2019 03:36:51

Diagnostics Check		OK	03:23	
Air Blank		0.000	03:23	
Control Test		0.079	03:24	
Air Blank		0.000	03:24	
Subject Sample	# 1	0.160	03:25	
Breath Volume		1.511	03:25	
Air Blank		0.000	03:25	
Air Blank		0.000	03:27	
Subject Sample	# 2	0.147	03:30	Volume Not Met
Breath Volume		0.789	03:30	
Air Blank		0.000	03:31	
Air Blank		0.000	03:33	
Subject Sample	# 3	0.158	03:35	
Breath Volume		1.500	03:35	
Air Blank		0.000	03:35	
Control Test		0.079	03:36	
Air Blank		0.000	03:36	
Diagnostics Check		OK	03:36	

Data Download Date/Time: 07/29/2019 15:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

INSTRUMENT REGISTERED TO: AUBURNDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALL,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Auburndale PD
ARRESTING OFFICER: WALL,CHRISTOPHER
ARRESTING OFFICER AGENCY: Auburndale PD
LAST AGENCY INSPECTION: 06/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: LYNN

DATE/TIME OF TEST: 07/01/2019 23:46:23

Diagnostics Check	OK	23:43	
Air Blank	0.000	23:44	
Control Test	0.080	23:44	
Air Blank	0.000	23:44	
Subject Sample	# 1	23:44	Subject Test Refused
Breath Volume	0.000	23:44	
Air Blank	0.000	23:45	
Control Test	0.080	23:45	
Air Blank	0.000	23:46	
Diagnostics Check	OK	23:46	

Data Download Date/Time: 07/26/2019 11:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

INSTRUMENT REGISTERED TO: AUBURNDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALL,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Auburndale PD
ARRESTING OFFICER: HAVRANEK,ADAM
ARRESTING OFFICER AGENCY: Auburndale PD
LAST AGENCY INSPECTION: 06/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 07/06/2019 03:36:59

Diagnostics Check	OK	03:29
Air Blank	0.000	03:30
Control Test	0.080	03:30
Air Blank	0.000	03:30
Subject Sample # 1	0.127	03:32
Breath Volume	1.757	03:32
Air Blank	0.000	03:32
Air Blank	0.000	03:34
Subject Sample # 2	0.128	03:35
Breath Volume	1.609	03:35
Air Blank	0.000	03:36
Control Test	0.079	03:36
Air Blank	0.000	03:36
Diagnostics Check	OK	03:36

Data Download Date/Time: 07/26/2019 11:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

INSTRUMENT REGISTERED TO: AUBURNDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAL,AUSTIN,J
BREATH TEST OPERATOR AGENCY: Auburndale PD
ARRESTING OFFICER: HAVRANEK,ADAM
ARRESTING OFFICER AGENCY: Auburndale PD
LAST AGENCY INSPECTION: 07/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 07/29/2019 22:55:52

Diagnostics Check	OK	22:52	
Air Blank	0.000	22:52	
Control Test	0.079	22:53	
Air Blank	0.000	22:53	
Subject Sample	# 1	22:54	Subject Test Refused
Breath Volume	0.000	22:54	
Air Blank	0.000	22:54	
Control Test	0.080	22:55	
Air Blank	0.000	22:55	
Diagnostics Check	OK	22:55	

Data Download Date/Time: 08/29/2019 15:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DONDE,J,
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: LOPEZ,N
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 07/22/2019 00:58:24

Diagnostics Check		OK	00:50
Air Blank		0.000	00:51
Control Test		0.080	00:51
Air Blank		0.000	00:51
Subject Sample	# 1	0.094	00:53
Breath Volume		1.359	00:53
Air Blank		0.000	00:54
Air Blank		0.000	00:56
Subject Sample	# 2	0.095	00:56
Breath Volume		1.414	00:56
Air Blank		0.000	00:57
Control Test		0.080	00:57
Air Blank		0.000	00:58
Diagnostics Check		OK	00:58

Data Download Date/Time: 08/05/2019 14:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BING, TRAE, J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HAMILTON, ROBERT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 07/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: BUCHER

DATE/TIME OF TEST: 07/24/2019 19:02:33

Diagnostics Check	OK	18:56
Air Blank	0.000	18:56
Control Test	0.077	18:56
Air Blank	0.000	18:57
Subject Sample # 1	0.061	18:57
Breath Volume	3.496	18:57
Air Blank	0.000	18:58
Air Blank	0.000	19:00
Subject Sample # 2	0.062	19:01
Breath Volume	1.390	19:01
Air Blank	0.000	19:01
Control Test	0.077	19:01
Air Blank	0.000	19:02
Diagnostics Check	OK	19:02

Data Download Date/Time: 07/25/2019 07:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES,STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HAMILTON,ROBERT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 07/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 07/25/2019 22:48:39

Diagnostics Check	OK	22:41
Air Blank	0.000	22:42
Control Test	0.076	22:42
Air Blank	0.000	22:43
Subject Sample # 1	0.000	22:44
Breath Volume	1.222	22:44
Air Blank	0.000	22:44
Air Blank	0.000	22:46
Subject Sample # 2	0.000	22:47
Breath Volume	1.316	22:47
Air Blank	0.000	22:47
Control Test	0.077	22:48
Air Blank	0.000	22:48
Diagnostics Check	OK	22:48

Data Download Date/Time: 08/26/2019 10:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUCAS,NATHANIEL,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LUCAS,NATHANIEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 07/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 07/26/2019 03:16:55

Diagnostics Check	OK	03:14	
Air Blank	0.000	03:14	
Control Test	0.076	03:14	
Air Blank	0.000	03:15	
Subject Sample	# 1	03:15	Subject Test Refused
Breath Volume	0.000	03:15	
Air Blank	0.000	03:16	
Control Test	0.076	03:16	
Air Blank	0.000	03:16	
Diagnostics Check	OK	03:16	

Data Download Date/Time: 08/26/2019 10:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALFRED, BRANDON, J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ALFRED, BRANDON
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 06:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:55

SUBJECT'S LAST NAME: BEATY 3RD

DATE/TIME OF TEST: 07/26/2019 10:23:22

Diagnostics Check	OK	10:16
Air Blank	0.000	10:16
Control Test	0.077	10:16
Air Blank	0.000	10:17
Subject Sample # 1	0.157	10:18
Breath Volume	3.011	10:18
Air Blank	0.000	10:19
Air Blank	0.000	10:21
Subject Sample # 2	0.161	10:21
Breath Volume	3.007	10:21
Air Blank	0.000	10:22
Control Test	0.076	10:22
Air Blank	0.000	10:23
Diagnostics Check	OK	10:23

Data Download Date/Time: 08/26/2019 10:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHRISTELLO,CLIFFORD,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 07/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:37

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 07/27/2019 19:07:33

Diagnostics Check	OK	18:58
Air Blank	0.000	18:59
Control Test	0.076	18:59
Air Blank	0.000	18:59
Subject Sample # 1	0.249	19:00
Breath Volume	2.054	19:00
Air Blank	0.000	19:00
Air Blank	0.000	19:02
Subject Sample # 2	0.240	19:05
Breath Volume	1.765	19:05
Air Blank	0.000	19:06
Control Test	0.078	19:06
Air Blank	0.000	19:07
Diagnostics Check	OK	19:07

Data Download Date/Time: 08/26/2019 10:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COKE,CLEVELAND,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 07/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: SEXTON

DATE/TIME OF TEST: 07/28/2019 02:03:47

Diagnostics Check	OK	02:00	
Air Blank	0.000	02:01	
Control Test	0.077	02:01	
Air Blank	0.000	02:01	
Subject Sample	# 1	02:02	Subject Test Refused
Breath Volume	0.000	02:02	
Air Blank	0.000	02:02	
Control Test	0.077	02:03	
Air Blank	0.000	02:03	
Diagnostics Check	OK	02:03	

Data Download Date/Time: 08/26/2019 10:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COKE,CLEVELAND,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 07/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1081808017
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: MATA

DATE/TIME OF TEST: 07/28/2019 03:55:33

Diagnostics Check		OK	03:45
Air Blank		0.000	03:45
Control Test		0.077	03:46
Air Blank		0.000	03:46
Subject Sample	# 1	0.198	03:47
Breath Volume		2.285	03:47
Air Blank		0.000	03:47
Air Blank		0.000	03:49
Subject Sample	# 2	0.175	03:50
Breath Volume		1.433	03:50
Air Blank		0.000	03:51
Air Blank		0.000	03:52
Subject Sample	# 3	0.170	03:53
Breath Volume		1.687	03:53
Air Blank		0.000	03:54
Control Test		0.075	03:54
Air Blank		0.000	03:55
Diagnostics Check		OK	03:55

Data Download Date/Time: 08/26/2019 10:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JEUDY,RINALDO,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HAMILTON,ROBERT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 07/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: ROACH

DATE/TIME OF TEST: 07/30/2019 02:03:30

Diagnostics Check	OK	02:00	
Air Blank	0.000	02:01	
Control Test	0.077	02:01	
Air Blank	0.000	02:01	
Subject Sample	# 1	02:02	Subject Test Refused
Breath Volume	0.000	02:02	
Air Blank	0.000	02:02	
Control Test	0.078	02:02	
Air Blank	0.000	02:03	
Diagnostics Check	OK	02:03	

Data Download Date/Time: 08/26/2019 10:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JEUDY,RINALDO,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: DURRANCE,JACOB
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 07/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: MACIAS

DATE/TIME OF TEST: 07/31/2019 04:10:06

Diagnostics Check	OK	04:03
Air Blank	0.000	04:03
Control Test	0.078	04:03
Air Blank	0.000	04:04
Subject Sample # 1	0.051	04:05
Breath Volume	1.843	04:05
Air Blank	0.000	04:05
Air Blank	0.000	04:07
Subject Sample # 2	0.054	04:08
Breath Volume	2.785	04:08
Air Blank	0.000	04:09
Control Test	0.078	04:09
Air Blank	0.000	04:10
Diagnostics Check	OK	04:10

Data Download Date/Time: 08/26/2019 10:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHRISTELLO,CLIFFORD,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: CHRISTELLO,CLIFFORD
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 07/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 12:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:59

SUBJECT'S LAST NAME: WOOSLEY

DATE/TIME OF TEST: 07/31/2019 13:26:40

Diagnostics Check	OK	13:19
Air Blank	0.000	13:19
Control Test	0.077	13:20
Air Blank	0.000	13:20
Subject Sample # 1	0.000	13:21
Breath Volume	1.453	13:21
Air Blank	0.000	13:21
Air Blank	0.000	13:23
Subject Sample # 2	0.000	13:25
Breath Volume	1.382	13:25
Air Blank	0.000	13:25
Control Test	0.076	13:26
Air Blank	0.000	13:26
Diagnostics Check	OK	13:26

Data Download Date/Time: 08/26/2019 10:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITTS,JAMES,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: OGG,DAVID
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 07/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:14

SUBJECT'S LAST NAME: RYKOWSKI

DATE/TIME OF TEST: 07/31/2019 17:53:44

Diagnostics Check		OK	17:46
Air Blank		0.000	17:47
Control Test		0.077	17:47
Air Blank		0.000	17:48
Subject Sample	# 1	0.219	17:48
Breath Volume		2.683	17:48
Air Blank		0.000	17:49
Air Blank		0.000	17:51
Subject Sample	# 2	0.212	17:52
Breath Volume		1.937	17:52
Air Blank		0.000	17:52
Control Test		0.076	17:53
Air Blank		0.000	17:53
Diagnostics Check		OK	17:53

Data Download Date/Time: 08/26/2019 10:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHESSHER,SANDRA,W
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CHESSHER,SANDRA
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 07/06/2019 02:25:15

Diagnostics Check	OK	02:18
Air Blank	0.000	02:19
Control Test	0.081	02:19
Air Blank	0.000	02:20
Subject Sample # 1	0.139	02:20
Breath Volume	3.445	02:20
Air Blank	0.000	02:21
Air Blank	0.000	02:23
Subject Sample # 2	0.142	02:23
Breath Volume	3.250	02:23
Air Blank	0.000	02:24
Control Test	0.081	02:24
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

Data Download Date/Time: 08/23/2019 16:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHESSHER,SANDRA,W
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CHESSHER,S
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: REED

DATE/TIME OF TEST: 07/06/2019 23:39:50

Diagnostics Check	OK	23:36	
Air Blank	0.000	23:37	
Control Test	0.082	23:37	
Air Blank	0.000	23:38	
Subject Sample	# 1	23:38	Subject Test Refused
Breath Volume	0.000	23:38	
Air Blank	0.000	23:38	
Control Test	0.083	23:39	
Air Blank	0.000	23:39	
Diagnostics Check	OK	23:39	

Data Download Date/Time: 08/23/2019 16:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHESSHER,SANDRA,W
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CHESSHER,SANDRA
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: LONG

DATE/TIME OF TEST: 07/07/2019 04:45:17

Diagnostics Check	OK	04:33	
Air Blank	0.000	04:33	
Control Test	0.078	04:33	
Air Blank	0.000	04:34	
Subject Sample	# 1 0.186	04:35	
Breath Volume	1.160	04:35	
Air Blank	0.000	04:35	
Air Blank	0.000	04:37	
Subject Sample	# 2 0.156	04:40	Volume Not Met
Breath Volume	0.199	04:40	
Air Blank	0.000	04:41	
Air Blank	0.000	04:43	
Subject Sample	# 3 0.184	04:43	
Breath Volume	1.535	04:43	
Air Blank	0.000	04:44	
Control Test	0.081	04:44	
Air Blank	0.000	04:45	
Diagnostics Check	OK	04:45	

Data Download Date/Time: 08/23/2019 16:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DAVID,B
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: BROWN,KRISTY
ARRESTING OFFICER AGENCY: Daytona Beach Shores
LAST AGENCY INSPECTION: 06/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: VON VOGT

DATE/TIME OF TEST: 07/01/2019 22:24:09

Diagnostics Check	OK	22:23	
Air Blank		22:23	Ambient Fail
Air Blank	0.000	22:24	

Data Download Date/Time: 07/20/2019 02:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DAVID,B
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: BROWN,KRISTY
ARRESTING OFFICER AGENCY: Daytona Beach Shores
LAST AGENCY INSPECTION: 06/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: VON VOGT

DATE/TIME OF TEST: 07/01/2019 22:54:17

Diagnostics Check	OK	22:48
Air Blank	0.000	22:48
Control Test	0.079	22:48
Air Blank	0.000	22:49
Subject Sample # 1	0.150	22:49
Breath Volume	2.093	22:49
Air Blank	0.000	22:50
Air Blank	0.000	22:52
Subject Sample # 2	0.143	22:52
Breath Volume	2.011	22:52
Air Blank	0.000	22:53
Control Test	0.078	22:53
Air Blank	0.000	22:54
Diagnostics Check	OK	22:54

Data Download Date/Time: 07/20/2019 02:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOTZ,KATHRYN,N
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: GOTZ,KATHRYN
ARRESTING OFFICER AGENCY: Daytona Beach Shores
LAST AGENCY INSPECTION: 06/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 07/05/2019 04:23:49

Diagnostics Check	OK	04:15
Air Blank	0.000	04:16
Control Test	0.080	04:16
Air Blank	0.000	04:17
Subject Sample # 1	0.188	04:19
Breath Volume	1.273	04:19
Air Blank	0.000	04:19
Air Blank	0.000	04:21
Subject Sample # 2	0.188	04:22
Breath Volume	2.367	04:22
Air Blank	0.000	04:22
Control Test	0.078	04:23
Air Blank	0.000	04:23
Diagnostics Check	OK	04:23

Data Download Date/Time: 07/20/2019 02:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOTZ,KATHRYN,N
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: CURTIS,CONNOR
ARRESTING OFFICER AGENCY: Daytona Beach Shores
LAST AGENCY INSPECTION: 07/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: MCREYNOLDS

DATE/TIME OF TEST: 07/28/2019 04:21:56

Diagnostics Check		OK	04:15
Air Blank		0.000	04:15
Control Test		0.080	04:16
Air Blank		0.000	04:16
Subject Sample	# 1	0.156	04:17
Breath Volume		2.246	04:17
Air Blank		0.000	04:18
Air Blank		0.000	04:19
Subject Sample	# 2	0.155	04:20
Breath Volume		2.003	04:20
Air Blank		0.000	04:21
Control Test		0.079	04:21
Air Blank		0.000	04:21
Diagnostics Check		OK	04:21

Data Download Date/Time: 08/30/2019 11:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ULESKI,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: RIERA,SCOTT
ARRESTING OFFICER AGENCY: New Smyrna Beach PD
LAST AGENCY INSPECTION: 07/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: TRAVIESO

DATE/TIME OF TEST: 07/30/2019 23:53:38

Diagnostics Check	OK	23:43
Air Blank	0.000	23:44
Control Test	0.079	23:44
Air Blank	0.000	23:45
Subject Sample # 1	0.254	23:45
Breath Volume	1.925	23:45
Air Blank	0.000	23:46
Air Blank	0.000	23:48
Subject Sample # 2	0.228	23:48
Breath Volume	1.964	23:48
Air Blank	0.000	23:49
Air Blank	0.000	23:51
Subject Sample # 3	0.226	23:52
Breath Volume	2.187	23:52
Air Blank	0.000	23:52
Control Test	0.078	23:53
Air Blank	0.000	23:53
Diagnostics Check	OK	23:53

Data Download Date/Time: 08/30/2019 11:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRESENZI,G,
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: POWERS,S
ARRESTING OFFICER AGENCY: OTHER Fed
LAST AGENCY INSPECTION: 06/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 765462
DRY GAS EXPIRY DATE: 04/13/2020
VIOLATION CODE: DUI
ARREST TIME: 19:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: BASS

DATE/TIME OF TEST: 07/20/2019 20:42:39

Diagnostics Check	OK	20:34	
Air Blank	0.000	20:35	
Control Test	0.075	20:35	
Air Blank	0.000	20:35	
Subject Sample	# 1 0.112	20:37	
Breath Volume	2.183	20:37	
Air Blank	0.000	20:38	
Air Blank	0.000	20:40	
Subject Sample	# 2 0.101	20:41	
Breath Volume	1.136	20:41	
Air Blank	0.000	20:41	
Control Test	0.072	20:42	Control Outside Tolerance
Air Blank	0.000	20:42	

Data Download Date/Time: 07/28/2019 11:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001244

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Putnam County SO
ARRESTING OFFICER: APPLEBY,THOMAS
ARRESTING OFFICER AGENCY: Putnam County SO
LAST AGENCY INSPECTION: 06/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A3
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: COOEY

DATE/TIME OF TEST: 07/04/2019 03:54:51

Diagnostics Check	OK	03:52	
Air Blank	0.000	03:52	
Control Test	0.080	03:52	
Air Blank	0.000	03:53	
Subject Sample	# 1	03:53	Subject Test Refused
Breath Volume	0.000	03:53	
Air Blank	0.000	03:53	
Control Test	0.080	03:54	
Air Blank	0.000	03:54	
Diagnostics Check	OK	03:54	

Data Download Date/Time: 07/22/2019 19:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001244

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WESTER,ALEXANDER,W
BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)
ARRESTING OFFICER: WESTER,ALEXANDER
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 06/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A3
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: BUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: SLAUGHTER

DATE/TIME OF TEST: 07/06/2019 22:05:41

Diagnostics Check	OK	21:59
Air Blank	0.000	22:00
Control Test	0.080	22:00
Air Blank	0.000	22:00
Subject Sample # 1	0.104	22:01
Breath Volume	3.187	22:01
Air Blank	0.000	22:01
Air Blank	0.000	22:03
Subject Sample # 2	0.100	22:04
Breath Volume	3.562	22:04
Air Blank	0.000	22:04
Control Test	0.079	22:05
Air Blank	0.000	22:05
Diagnostics Check	OK	22:05

Data Download Date/Time: 07/22/2019 19:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001244

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROUNTREE,JUSTIN,T
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: ROUNTREE,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A3
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: HAEN

DATE/TIME OF TEST: 07/07/2019 05:46:30

Diagnostics Check		OK	05:37
Air Blank		0.000	05:38
Control Test		0.080	05:38
Air Blank		0.000	05:39
Subject Sample	# 1	0.216	05:41
Breath Volume		1.671	05:41
Air Blank		0.000	05:42
Air Blank		0.000	05:44
Subject Sample	# 2	0.219	05:44
Breath Volume		2.070	05:44
Air Blank		0.000	05:45
Control Test		0.080	05:45
Air Blank		0.000	05:46
Diagnostics Check		OK	05:46

Data Download Date/Time: 07/22/2019 19:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001244

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS, TIFFANY, R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: WILLIAMS, TIFFANY
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 07/22/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 29017080A3
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: ARDOLINO

DATE/TIME OF TEST: 07/29/2019 00:31:12

Diagnostics Check	OK	00:23
Air Blank	0.000	00:24
Control Test	0.081	00:24
Air Blank	0.000	00:24
Subject Sample # 1	0.150	00:26
Breath Volume	2.386	00:26
Air Blank	0.000	00:27
Air Blank	0.000	00:29
Subject Sample # 2	0.151	00:29
Breath Volume	1.949	00:29
Air Blank	0.000	00:30
Control Test	0.080	00:30
Air Blank	0.000	00:31
Diagnostics Check	OK	00:31

Data Download Date/Time: 08/20/2019 18:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001246

INSTRUMENT REGISTERED TO: SOUTH DAYTONA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIDONATO,MIKE,J
BREATH TEST OPERATOR AGENCY: South Daytona PD
ARRESTING OFFICER: BAZE,KAYLOR
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 06/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: FAGAN

DATE/TIME OF TEST: 07/28/2019 04:10:29

Diagnostics Check	OK	04:04
Air Blank	0.000	04:04
Control Test	0.081	04:04
Air Blank	0.000	04:05
Subject Sample # 1	0.084	04:05
Breath Volume	1.785	04:05
Air Blank	0.000	04:06
Air Blank	0.000	04:08
Subject Sample # 2	0.083	04:09
Breath Volume	2.085	04:09
Air Blank	0.000	04:09
Control Test	0.082	04:09
Air Blank	0.000	04:10
Diagnostics Check	OK	04:10

Data Download Date/Time: 07/31/2019 10:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBIN,GASTON,A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: ROBIN,GASTON
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A3
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: DANG

DATE/TIME OF TEST: 07/09/2019 02:40:18

Diagnostics Check	OK	02:34
Air Blank	0.000	02:34
Control Test	0.080	02:34
Air Blank	0.000	02:35
Subject Sample # 1	0.121	02:35
Breath Volume	1.449	02:35
Air Blank	0.000	02:36
Air Blank	0.000	02:38
Subject Sample # 2	0.117	02:38
Breath Volume	1.343	02:38
Air Blank	0.000	02:39
Control Test	0.080	02:39
Air Blank	0.000	02:40
Diagnostics Check	OK	02:40

Data Download Date/Time: 07/28/2019 22:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRATSCH,CASEY,R
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: SHINAULT,RYAN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/28/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 19817080A3
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 10:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:05

SUBJECT'S LAST NAME: HULL

DATE/TIME OF TEST: 07/09/2019 11:31:55

Diagnostics Check	OK	11:25
Air Blank	0.000	11:26
Control Test	0.081	11:26
Air Blank	0.000	11:27
Subject Sample # 1	0.000	11:27
Breath Volume	1.457	11:27
Air Blank	0.000	11:27
Air Blank	0.000	11:29
Subject Sample # 2	0.000	11:30
Breath Volume	1.550	11:30
Air Blank	0.000	11:31
Control Test	0.080	11:31
Air Blank	0.000	11:31
Diagnostics Check	OK	11:31

Data Download Date/Time: 07/28/2019 22:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HALLOCK,STEPHEN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 07/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: DELAROSAMORALES

DATE/TIME OF TEST: 07/25/2019 02:45:57

Diagnostics Check	OK	02:38
Air Blank	0.000	02:39
Control Test	0.080	02:39
Air Blank	0.000	02:39
Subject Sample # 1	0.152	02:41
Breath Volume	2.316	02:41
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.146	02:44
Breath Volume	2.933	02:44
Air Blank	0.000	02:45
Control Test	0.080	02:45
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 07/31/2019 15:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,ARMAND,G
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AUSTIN,ADAM
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 07/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:38

SUBJECT'S LAST NAME: WHALAN

DATE/TIME OF TEST: 07/25/2019 16:10:23

Diagnostics Check	OK	16:04
Air Blank	0.000	16:04
Control Test	0.081	16:04
Air Blank	0.000	16:05
Subject Sample # 1	0.000	16:06
Breath Volume	2.886	16:06
Air Blank	0.000	16:06
Air Blank	0.000	16:08
Subject Sample # 2	0.000	16:09
Breath Volume	2.179	16:09
Air Blank	0.000	16:09
Control Test	0.080	16:09
Air Blank	0.000	16:10
Diagnostics Check	OK	16:10

Data Download Date/Time: 07/31/2019 15:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HALLOCK,STEPHEN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 07/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: HERNANDEZ CRUZ

DATE/TIME OF TEST: 07/26/2019 02:10:41

Diagnostics Check		OK	01:58	
Air Blank		0.000	01:59	
Control Test		0.081	01:59	
Air Blank		0.000	02:00	
Subject Sample	# 1	0.168	02:03	Volume Not Met
Breath Volume		0.300	02:03	
Air Blank		0.000	02:04	
Air Blank		0.000	02:05	
Subject Sample	# 2	0.174	02:08	Volume Not Met
Breath Volume		0.195	02:08	
Air Blank		0.000	02:09	
Control Test		0.079	02:10	
Air Blank		0.000	02:10	
Diagnostics Check		OK	02:10	

Data Download Date/Time: 07/31/2019 15:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,ARMAND,G
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WELSCHINGER,MICKAEL
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:38

SUBJECT'S LAST NAME: OHEARN

DATE/TIME OF TEST: 07/26/2019 16:09:04

Diagnostics Check	OK	16:02
Air Blank	0.000	16:03
Control Test	0.081	16:03
Air Blank	0.000	16:04
Subject Sample # 1	0.146	16:04
Breath Volume	1.636	16:04
Air Blank	0.000	16:05
Air Blank	0.000	16:07
Subject Sample # 2	0.149	16:07
Breath Volume	1.386	16:07
Air Blank	0.000	16:08
Control Test	0.080	16:08
Air Blank	0.000	16:08
Diagnostics Check	OK	16:09

Data Download Date/Time: 07/31/2019 15:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VENTURY,RONNIE
ARRESTING OFFICER AGENCY: Edgewood PD
LAST AGENCY INSPECTION: 07/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:01

SUBJECT'S LAST NAME: RIEDERICH

DATE/TIME OF TEST: 07/26/2019 23:32:17

Diagnostics Check	OK	23:25
Air Blank	0.000	23:25
Control Test	0.082	23:25
Air Blank	0.000	23:26
Subject Sample # 1	0.198	23:26
Breath Volume	2.363	23:26
Air Blank	0.000	23:27
Air Blank	0.000	23:29
Subject Sample # 2	0.191	23:30
Breath Volume	1.152	23:30
Air Blank	0.000	23:31
Control Test	0.081	23:31
Air Blank	0.000	23:32
Diagnostics Check	OK	23:32

Data Download Date/Time: 07/31/2019 15:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BELFORD,DAVID
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: DETWILER

DATE/TIME OF TEST: 07/27/2019 00:51:09

Diagnostics Check	OK	00:44
Air Blank	0.000	00:44
Control Test	0.082	00:45
Air Blank	0.000	00:45
Subject Sample # 1	0.178	00:46
Breath Volume	4.726	00:46
Air Blank	0.000	00:47
Air Blank	0.000	00:48
Subject Sample # 2	0.171	00:49
Breath Volume	3.480	00:49
Air Blank	0.000	00:50
Control Test	0.080	00:50
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

Data Download Date/Time: 07/31/2019 15:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MARTINEZ,YAISEL
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: SABAU

DATE/TIME OF TEST: 07/27/2019 01:52:51

Diagnostics Check	OK	01:45
Air Blank	0.000	01:46
Control Test	0.081	01:46
Air Blank	0.000	01:46
Subject Sample # 1	0.177	01:48
Breath Volume	3.101	01:48
Air Blank	0.000	01:49
Air Blank	0.000	01:50
Subject Sample # 2	0.168	01:51
Breath Volume	2.527	01:51
Air Blank	0.000	01:51
Control Test	0.079	01:52
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

Data Download Date/Time: 07/31/2019 15:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: IRELAND,DAVID
ARRESTING OFFICER AGENCY: Edgewood PD
LAST AGENCY INSPECTION: 07/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: OQUENDO ALVARADO

DATE/TIME OF TEST: 07/27/2019 02:56:35

Diagnostics Check	OK	02:49
Air Blank	0.000	02:50
Control Test	0.081	02:50
Air Blank	0.000	02:51
Subject Sample # 1	0.091	02:52
Breath Volume	1.570	02:52
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample # 2	0.088	02:55
Breath Volume	1.359	02:55
Air Blank	0.000	02:55
Control Test	0.079	02:55
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 07/31/2019 15:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LANGHORST, BRYAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 07/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 06:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:02

SUBJECT'S LAST NAME: GOLDMAN

DATE/TIME OF TEST: 07/27/2019 07:45:49

Diagnostics Check	OK	07:39
Air Blank	0.000	07:40
Control Test	0.081	07:40
Air Blank	0.000	07:40
Subject Sample # 1	0.232	07:41
Breath Volume	2.140	07:41
Air Blank	0.000	07:41
Air Blank	0.000	07:43
Subject Sample # 2	0.229	07:44
Breath Volume	2.816	07:44
Air Blank	0.000	07:44
Control Test	0.081	07:45
Air Blank	0.000	07:45
Diagnostics Check	OK	07:45

Data Download Date/Time: 07/31/2019 15:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CRESSWELL,FRANK
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 07/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:16

SUBJECT'S LAST NAME: FICCA

DATE/TIME OF TEST: 07/27/2019 22:46:59

Diagnostics Check	OK	22:40
Air Blank	0.000	22:41
Control Test	0.081	22:41
Air Blank	0.000	22:42
Subject Sample # 1	0.071	22:42
Breath Volume	3.578	22:42
Air Blank	0.000	22:43
Air Blank	0.000	22:45
Subject Sample # 2	0.061	22:45
Breath Volume	1.816	22:45
Air Blank	0.000	22:46
Control Test	0.080	22:46
Air Blank	0.000	22:46
Diagnostics Check	OK	22:46

Data Download Date/Time: 07/31/2019 15:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JUNCO,JONATHAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 07/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 07/28/2019 01:45:38

Diagnostics Check		OK	01:39
Air Blank		0.000	01:39
Control Test		0.081	01:39
Air Blank		0.000	01:40
Subject Sample	# 1	0.223	01:40
Breath Volume		2.210	01:40
Air Blank		0.000	01:41
Air Blank		0.000	01:43
Subject Sample	# 2	0.215	01:43
Breath Volume		2.140	01:43
Air Blank		0.000	01:44
Control Test		0.081	01:45
Air Blank		0.000	01:45
Diagnostics Check		OK	01:45

Data Download Date/Time: 07/31/2019 15:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ABREU,JUAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 07/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 07/28/2019 03:51:14

Diagnostics Check	OK	03:45
Air Blank	0.000	03:45
Control Test	0.081	03:45
Air Blank	0.000	03:46
Subject Sample # 1	0.309	03:46
Breath Volume	2.875	03:46
Air Blank	0.000	03:47
Air Blank	0.000	03:49
Subject Sample # 2	0.308	03:49
Breath Volume	2.781	03:49
Air Blank	0.000	03:50
Control Test	0.080	03:50
Air Blank	0.000	03:51
Diagnostics Check	OK	03:51

Data Download Date/Time: 07/31/2019 15:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JORGE, JUNIOR
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 07/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: LALBAHADUR

DATE/TIME OF TEST: 07/28/2019 04:03:40

Diagnostics Check		OK	03:56
Air Blank		0.000	03:57
Control Test		0.081	03:57
Air Blank		0.000	03:58
Subject Sample	# 1	0.139	03:58
Breath Volume		1.796	03:58
Air Blank		0.000	03:59
Air Blank		0.000	04:01
Subject Sample	# 2	0.135	04:02
Breath Volume		1.242	04:02
Air Blank		0.000	04:02
Control Test		0.080	04:03
Air Blank		0.000	04:03
Diagnostics Check		OK	04:03

Data Download Date/Time: 07/31/2019 15:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: POWELL,KRISTI
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: SOLACHE

DATE/TIME OF TEST: 07/28/2019 05:21:11

Diagnostics Check	OK	05:13
Air Blank	0.000	05:14
Control Test	0.081	05:14
Air Blank	0.000	05:15
Subject Sample # 1	0.091	05:16
Breath Volume	1.730	05:16
Air Blank	0.000	05:17
Air Blank	0.000	05:19
Subject Sample # 2	0.090	05:19
Breath Volume	1.164	05:19
Air Blank	0.000	05:20
Control Test	0.079	05:20
Air Blank	0.000	05:21
Diagnostics Check	OK	05:21

Data Download Date/Time: 07/31/2019 15:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RIVERA, LIBORIO
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 07/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 06:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:52

SUBJECT'S LAST NAME: DE LA ROSA

DATE/TIME OF TEST: 07/28/2019 07:19:37

Diagnostics Check		OK	07:13
Air Blank		0.000	07:13
Control Test		0.082	07:14
Air Blank		0.000	07:14
Subject Sample	# 1	0.201	07:15
Breath Volume		1.253	07:15
Air Blank		0.000	07:15
Air Blank		0.000	07:17
Subject Sample	# 2	0.197	07:17
Breath Volume		2.000	07:17
Air Blank		0.000	07:18
Control Test		0.081	07:19
Air Blank		0.000	07:19
Diagnostics Check		OK	07:19

Data Download Date/Time: 07/31/2019 15:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HARMAN,RICHARD
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: PEREZ DIAZ

DATE/TIME OF TEST: 07/28/2019 21:55:07

Diagnostics Check	OK	21:47
Air Blank	0.000	21:48
Control Test	0.081	21:48
Air Blank	0.000	21:48
Subject Sample # 1	0.217	21:50
Breath Volume	3.773	21:50
Air Blank	0.000	21:51
Air Blank	0.000	21:52
Subject Sample # 2	0.204	21:53
Breath Volume	2.652	21:53
Air Blank	0.000	21:54
Control Test	0.078	21:54
Air Blank	0.000	21:55
Diagnostics Check	OK	21:55

Data Download Date/Time: 07/31/2019 15:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JUNCO,JONATHAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 07/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: MOTALVESDOSSANTOS

DATE/TIME OF TEST: 07/28/2019 22:39:45

Diagnostics Check	OK	22:31
Air Blank	0.000	22:32
Control Test	0.080	22:32
Air Blank	0.000	22:32
Subject Sample # 1	0.232	22:35
Breath Volume	1.332	22:35
Air Blank	0.000	22:35
Air Blank	0.000	22:37
Subject Sample # 2	0.239	22:38
Breath Volume	1.570	22:38
Air Blank	0.000	22:38
Control Test	0.078	22:39
Air Blank	0.000	22:39
Diagnostics Check	OK	22:39

Data Download Date/Time: 07/31/2019 15:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELVILLE,KELLY,M
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TUERO,ADRIAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 09:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:43

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 07/29/2019 10:11:24

Diagnostics Check	Fail	10:10	Voltage/Current Test
Air Blank	0.000	10:11	

Data Download Date/Time: 07/31/2019 15:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KERN,ETHAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 06/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: ORTA

DATE/TIME OF TEST: 07/02/2019 04:05:03

Diagnostics Check	OK	03:59
Air Blank	0.000	03:59
Control Test	0.080	03:59
Air Blank	0.000	04:00
Subject Sample # 1	0.062	04:00
Breath Volume	1.828	04:00
Air Blank	0.000	04:01
Air Blank	0.000	04:03
Subject Sample # 2	0.060	04:03
Breath Volume	2.003	04:03
Air Blank	0.000	04:04
Control Test	0.080	04:04
Air Blank	0.000	04:04
Diagnostics Check	OK	04:05

Data Download Date/Time: 07/04/2019 16:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BINET,MARLON
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 06/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 07/03/2019 03:06:10

Diagnostics Check	OK	02:58
Air Blank	0.000	02:58
Control Test	0.079	02:58
Air Blank	0.000	02:59
Subject Sample # 1	0.227	03:01
Breath Volume	1.386	03:01
Air Blank	0.000	03:02
Air Blank	0.000	03:03
Subject Sample # 2	0.231	03:04
Breath Volume	1.277	03:04
Air Blank	0.000	03:05
Control Test	0.078	03:05
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

Data Download Date/Time: 07/04/2019 16:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MARTELL,SIONE
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 06/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: HALLAL

DATE/TIME OF TEST: 07/04/2019 00:41:57

Diagnostics Check	OK	00:36
Air Blank	0.000	00:36
Control Test	0.080	00:36
Air Blank	0.000	00:37
Subject Sample # 1	0.000	00:37
Breath Volume	3.707	00:37
Air Blank	0.000	00:38
Air Blank	0.000	00:40
Subject Sample # 2	0.000	00:40
Breath Volume	3.644	00:40
Air Blank	0.000	00:41
Control Test	0.079	00:41
Air Blank	0.000	00:41
Diagnostics Check	OK	00:41

Data Download Date/Time: 07/04/2019 16:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SIMPSON,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: LAURY

DATE/TIME OF TEST: 07/04/2019 01:16:16

Diagnostics Check	OK	01:10
Air Blank	0.000	01:10
Control Test	0.080	01:10
Air Blank	0.000	01:11
Subject Sample # 1	0.136	01:11
Breath Volume	1.507	01:11
Air Blank	0.000	01:12
Air Blank	0.000	01:14
Subject Sample # 2	0.146	01:14
Breath Volume	1.789	01:14
Air Blank	0.000	01:15
Control Test	0.079	01:15
Air Blank	0.000	01:16
Diagnostics Check	OK	01:16

Data Download Date/Time: 07/04/2019 16:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LEE,MIKEL
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 06/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: BONINO

DATE/TIME OF TEST: 07/04/2019 05:03:25

Diagnostics Check		OK	04:56
Air Blank		0.000	04:57
Control Test		0.080	04:57
Air Blank		0.000	04:57
Subject Sample	# 1	0.134	04:58
Breath Volume		1.710	04:58
Air Blank		0.000	04:59
Air Blank		0.000	05:01
Subject Sample	# 2	0.118	05:01
Breath Volume		2.148	05:01
Air Blank		0.000	05:02
Control Test		0.079	05:02
Air Blank		0.000	05:03
Diagnostics Check		OK	05:03

Data Download Date/Time: 07/04/2019 16:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,ARMAND,G
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AUSTIN,ADAM
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 06/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:15

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/04/2019 16:49:23

Diagnostics Check	OK	16:43
Air Blank	0.000	16:43
Control Test	0.080	16:43
Air Blank	0.000	16:44
Subject Sample # 1	0.000	16:45
Breath Volume	2.476	16:45
Air Blank	0.000	16:45
Air Blank	0.000	16:47
Subject Sample # 2	0.000	16:48
Breath Volume	3.132	16:48
Air Blank	0.000	16:48
Control Test	0.079	16:48
Air Blank	0.000	16:49
Diagnostics Check	OK	16:49

Data Download Date/Time: 07/11/2019 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VENTURA,RONNIE
ARRESTING OFFICER AGENCY: Edgewood PD
LAST AGENCY INSPECTION: 06/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: FAIRCLOTH

DATE/TIME OF TEST: 07/05/2019 02:30:10

Diagnostics Check	OK	02:23
Air Blank	0.000	02:24
Control Test	0.079	02:24
Air Blank	0.000	02:25
Subject Sample # 1	0.208	02:25
Breath Volume	2.679	02:25
Air Blank	0.000	02:26
Air Blank	0.000	02:28
Subject Sample # 2	0.193	02:28
Breath Volume	2.859	02:28
Air Blank	0.000	02:29
Control Test	0.079	02:29
Air Blank	0.000	02:30
Diagnostics Check	OK	02:30

Data Download Date/Time: 07/11/2019 16:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RODRIGUEZ,STEPHANIE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 06/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: VELAZQUEZ

DATE/TIME OF TEST: 07/05/2019 03:47:52

Diagnostics Check	OK	03:36	
Air Blank	0.000	03:36	
Control Test	0.079	03:37	
Air Blank	0.000	03:37	
Subject Sample	# 1 0.052	03:40	Volume Not Met
Breath Volume	0.078	03:40	
Air Blank	0.000	03:41	
Air Blank	0.000	03:43	
Subject Sample	# 2 0.054	03:46	Volume Not Met
Breath Volume	0.121	03:46	
Air Blank	0.000	03:46	
Control Test	0.079	03:47	
Air Blank	0.000	03:47	
Diagnostics Check	OK	03:47	

Data Download Date/Time: 07/11/2019 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAWKINS,DEB
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: SQUILLANTE

DATE/TIME OF TEST: 07/05/2019 04:13:35

Diagnostics Check	OK	04:07
Air Blank	0.000	04:07
Control Test	0.078	04:08
Air Blank	0.000	04:08
Subject Sample # 1	0.158	04:09
Breath Volume	1.992	04:09
Air Blank	0.000	04:09
Air Blank	0.000	04:11
Subject Sample # 2	0.157	04:12
Breath Volume	1.781	04:12
Air Blank	0.000	04:12
Control Test	0.079	04:13
Air Blank	0.000	04:13
Diagnostics Check	OK	04:13

Data Download Date/Time: 07/11/2019 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MICHAEL, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: THATHURIN, W
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 06/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 09:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:50

SUBJECT'S LAST NAME: GARY

DATE/TIME OF TEST: 07/05/2019 10:25:26

Diagnostics Check	OK	10:17
Air Blank	0.000	10:17
Control Test	0.079	10:17
Air Blank	0.000	10:18
Subject Sample # 1	0.134	10:19
Breath Volume	1.253	10:19
Air Blank	0.000	10:20
Air Blank	0.000	10:21
Subject Sample # 2	0.131	10:23
Breath Volume	1.585	10:23
Air Blank	0.000	10:24
Control Test	0.079	10:24
Air Blank	0.000	10:25
Diagnostics Check	OK	10:25

Data Download Date/Time: 07/11/2019 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MICHAEL, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MCLEAN, NICHOLAS
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 14:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:45

SUBJECT'S LAST NAME: HANIFF

DATE/TIME OF TEST: 07/05/2019 16:17:11

Diagnostics Check	OK	16:10
Air Blank	0.000	16:11
Control Test	0.080	16:11
Air Blank	0.000	16:12
Subject Sample # 1	0.000	16:12
Breath Volume	1.781	16:12
Air Blank	0.000	16:13
Air Blank	0.000	16:15
Subject Sample # 2	0.000	16:15
Breath Volume	2.558	16:15
Air Blank	0.000	16:16
Control Test	0.079	16:16
Air Blank	0.000	16:17
Diagnostics Check	OK	16:17

Data Download Date/Time: 07/11/2019 16:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CAMPBELL,ROBERT
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 06/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: OCASIO

DATE/TIME OF TEST: 07/05/2019 23:48:06

Diagnostics Check	OK	23:42
Air Blank	0.000	23:42
Control Test	0.079	23:43
Air Blank	0.000	23:43
Subject Sample # 1	0.000	23:43
Breath Volume	1.625	23:43
Air Blank	0.000	23:44
Air Blank	0.000	23:46
Subject Sample # 2	0.000	23:46
Breath Volume	1.667	23:46
Air Blank	0.000	23:47
Control Test	0.079	23:47
Air Blank	0.000	23:48
Diagnostics Check	OK	23:48

Data Download Date/Time: 07/11/2019 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MENDOZA,ANGEL
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 06/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: GORDON

DATE/TIME OF TEST: 07/06/2019 00:19:47

Diagnostics Check	OK	00:08	
Air Blank	0.000	00:09	
Control Test	0.079	00:09	
Air Blank	0.000	00:09	
Subject Sample	# 1	00:13	No Sample Provided
Breath Volume	0.000	00:13	
Air Blank	0.000	00:13	
Air Blank	0.000	00:15	
Subject Sample	# 2	00:18	Subject Test Refused
Breath Volume	0.000	00:18	
Air Blank	0.000	00:18	
Control Test	0.079	00:19	
Air Blank	0.000	00:19	
Diagnostics Check	OK	00:19	

Data Download Date/Time: 07/11/2019 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ANDREWS,GARY
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 06/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: ABEAR

DATE/TIME OF TEST: 07/06/2019 01:42:18

Diagnostics Check	OK	01:34
Air Blank	0.000	01:35
Control Test	0.080	01:35
Air Blank	0.000	01:36
Subject Sample # 1	0.119	01:37
Breath Volume	1.140	01:37
Air Blank	0.000	01:38
Air Blank	0.000	01:39
Subject Sample # 2	0.122	01:40
Breath Volume	1.699	01:40
Air Blank	0.000	01:41
Control Test	0.078	01:41
Air Blank	0.000	01:42
Diagnostics Check	OK	01:42

Data Download Date/Time: 07/11/2019 16:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BINET,MARLON
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 06/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 07/06/2019 03:15:56

Diagnostics Check	OK	03:10
Air Blank	0.000	03:10
Control Test	0.079	03:10
Air Blank	0.000	03:11
Subject Sample # 1	0.000	03:11
Breath Volume	1.621	03:11
Air Blank	0.000	03:12
Air Blank	0.000	03:14
Subject Sample # 2	0.000	03:14
Breath Volume	1.660	03:14
Air Blank	0.000	03:15
Control Test	0.080	03:15
Air Blank	0.000	03:15
Diagnostics Check	OK	03:15

Data Download Date/Time: 07/11/2019 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HARRIS,BRIAN
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 06/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: GUTIERREZ

DATE/TIME OF TEST: 07/06/2019 03:42:01

Diagnostics Check	OK	03:35
Air Blank	0.000	03:36
Control Test	0.079	03:36
Air Blank	0.000	03:37
Subject Sample # 1	0.120	03:37
Breath Volume	1.375	03:37
Air Blank	0.000	03:38
Air Blank	0.000	03:40
Subject Sample # 2	0.119	03:40
Breath Volume	1.417	03:40
Air Blank	0.000	03:41
Control Test	0.079	03:41
Air Blank	0.000	03:41
Diagnostics Check	OK	03:42

Data Download Date/Time: 07/11/2019 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HARMAN,RICHARD
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: MASUCCI

DATE/TIME OF TEST: 07/07/2019 02:59:06

Diagnostics Check	OK	02:52
Air Blank	0.000	02:53
Control Test	0.080	02:53
Air Blank	0.000	02:54
Subject Sample # 1	0.065	02:54
Breath Volume	2.375	02:54
Air Blank	0.000	02:55
Air Blank	0.000	02:57
Subject Sample # 2	0.065	02:57
Breath Volume	2.480	02:57
Air Blank	0.000	02:58
Control Test	0.079	02:58
Air Blank	0.000	02:59
Diagnostics Check	OK	02:59

Data Download Date/Time: 07/11/2019 16:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PARRISH,KAYLA
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 06/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 07/07/2019 03:39:56

Diagnostics Check	OK	03:33
Air Blank	0.000	03:34
Control Test	0.080	03:34
Air Blank	0.000	03:34
Subject Sample # 1	0.115	03:35
Breath Volume	2.398	03:35
Air Blank	0.000	03:35
Air Blank	0.000	03:37
Subject Sample # 2	0.113	03:38
Breath Volume	2.753	03:38
Air Blank	0.000	03:39
Control Test	0.079	03:39
Air Blank	0.000	03:39
Diagnostics Check	OK	03:39

Data Download Date/Time: 07/11/2019 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BINET, MARLON
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 06/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 06:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:25

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 07/07/2019 06:54:04

Diagnostics Check	OK	06:47
Air Blank	0.000	06:48
Control Test	0.081	06:48
Air Blank	0.000	06:49
Subject Sample # 1	0.000	06:49
Breath Volume	2.175	06:49
Air Blank	0.000	06:50
Air Blank	0.000	06:52
Subject Sample # 2	0.000	06:52
Breath Volume	2.578	06:52
Air Blank	0.000	06:53
Control Test	0.080	06:53
Air Blank	0.000	06:53
Diagnostics Check	OK	06:54

Data Download Date/Time: 07/11/2019 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PRICE,FORREST
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 06/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:21

SUBJECT'S LAST NAME: ODELL

DATE/TIME OF TEST: 07/07/2019 20:52:25

Diagnostics Check	OK	20:45
Air Blank	0.000	20:46
Control Test	0.080	20:46
Air Blank	0.000	20:47
Subject Sample # 1	0.240	20:47
Breath Volume	1.281	20:47
Air Blank	0.000	20:48
Air Blank	0.000	20:50
Subject Sample # 2	0.247	20:50
Breath Volume	2.164	20:50
Air Blank	0.000	20:51
Control Test	0.079	20:51
Air Blank	0.000	20:52
Diagnostics Check	OK	20:52

Data Download Date/Time: 07/11/2019 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MILLER,JACOB
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 06/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: PERDOMOHERNANDEZ

DATE/TIME OF TEST: 07/07/2019 22:46:34

Diagnostics Check	OK	22:39
Air Blank	0.000	22:40
Control Test	0.080	22:40
Air Blank	0.000	22:41
Subject Sample # 1	0.098	22:41
Breath Volume	1.250	22:41
Air Blank	0.000	22:42
Air Blank	0.000	22:44
Subject Sample # 2	0.097	22:45
Breath Volume	2.750	22:45
Air Blank	0.000	22:45
Control Test	0.079	22:45
Air Blank	0.000	22:46
Diagnostics Check	OK	22:46

Data Download Date/Time: 07/11/2019 16:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KNIGHT,RYAN
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 06/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 07/10/2019 00:33:35

Diagnostics Check	OK	00:27
Air Blank	0.000	00:27
Control Test	0.080	00:28
Air Blank	0.000	00:28
Subject Sample # 1	0.000	00:28
Breath Volume	1.312	00:28
Air Blank	0.000	00:29
Air Blank	0.000	00:31
Subject Sample # 2	0.000	00:32
Breath Volume	2.121	00:32
Air Blank	0.000	00:32
Control Test	0.080	00:33
Air Blank	0.000	00:33
Diagnostics Check	OK	00:33

Data Download Date/Time: 07/11/2019 16:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MENDOZA,ANGEL
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 06/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: FLETCHER

DATE/TIME OF TEST: 07/10/2019 01:42:02

Diagnostics Check	OK	01:35
Air Blank	0.000	01:36
Control Test	0.080	01:36
Air Blank	0.000	01:37
Subject Sample # 1	0.155	01:37
Breath Volume	3.195	01:37
Air Blank	0.000	01:38
Air Blank	0.000	01:39
Subject Sample # 2	0.157	01:40
Breath Volume	3.601	01:40
Air Blank	0.000	01:41
Control Test	0.079	01:41
Air Blank	0.000	01:41
Diagnostics Check	OK	01:42

Data Download Date/Time: 07/11/2019 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ALLEN,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: STOVER

DATE/TIME OF TEST: 07/10/2019 03:56:00

Diagnostics Check	OK	03:49
Air Blank	0.000	03:50
Control Test	0.080	03:50
Air Blank	0.000	03:51
Subject Sample # 1	0.000	03:51
Breath Volume	1.460	03:51
Air Blank	0.000	03:52
Air Blank	0.000	03:54
Subject Sample # 2	0.000	03:54
Breath Volume	1.476	03:54
Air Blank	0.000	03:55
Control Test	0.080	03:55
Air Blank	0.000	03:55
Diagnostics Check	OK	03:56

Data Download Date/Time: 07/11/2019 16:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MOYE, JAMES
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 06/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: HERNANDEZ LOPEZ

DATE/TIME OF TEST: 07/10/2019 04:24:56

Diagnostics Check	OK	04:11	
Air Blank	0.000	04:11	
Control Test	0.080	04:12	
Air Blank	0.000	04:12	
Subject Sample	# 1	04:15	No Sample Provided
Breath Volume	0.000	04:15	
Air Blank	0.000	04:16	
Air Blank	0.000	04:18	
Subject Sample	# 2	0.123	04:19
Breath Volume	2.136	04:19	
Air Blank	0.000	04:19	
Air Blank	0.000	04:21	
Subject Sample	# 3	0.131	04:23
Breath Volume	2.273	04:23	
Air Blank	0.000	04:24	
Control Test	0.079	04:24	
Air Blank	0.000	04:24	
Diagnostics Check	OK	04:24	

Data Download Date/Time: 07/11/2019 16:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HARWOOD,ARMANDO
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 06/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: CANDELL

DATE/TIME OF TEST: 07/12/2019 03:48:48

Diagnostics Check	OK	03:42
Air Blank	0.000	03:43
Control Test	0.080	03:43
Air Blank	0.000	03:43
Subject Sample # 1	0.101	03:44
Breath Volume	1.246	03:44
Air Blank	0.000	03:44
Air Blank	0.000	03:46
Subject Sample # 2	0.102	03:47
Breath Volume	2.578	03:47
Air Blank	0.000	03:47
Control Test	0.080	03:48
Air Blank	0.000	03:48
Diagnostics Check	OK	03:48

Data Download Date/Time: 07/18/2019 15:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SEVERE,JEFF
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 06/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: MARINO

DATE/TIME OF TEST: 07/13/2019 02:50:26

Diagnostics Check	OK	02:44
Air Blank	0.000	02:44
Control Test	0.080	02:45
Air Blank	0.000	02:45
Subject Sample # 1	0.176	02:46
Breath Volume	2.054	02:46
Air Blank	0.000	02:46
Air Blank	0.000	02:48
Subject Sample # 2	0.172	02:48
Breath Volume	1.902	02:48
Air Blank	0.000	02:49
Control Test	0.079	02:49
Air Blank	0.000	02:50
Diagnostics Check	OK	02:50

Data Download Date/Time: 07/18/2019 15:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SONS,KARA
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 06/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 07:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:07

SUBJECT'S LAST NAME: CASHES

DATE/TIME OF TEST: 07/13/2019 08:38:23

Diagnostics Check	OK	08:31
Air Blank	0.000	08:32
Control Test	0.079	08:32
Air Blank	0.000	08:33
Subject Sample # 1	0.198	08:33
Breath Volume	3.750	08:33
Air Blank	0.000	08:34
Air Blank	0.000	08:36
Subject Sample # 2	0.202	08:36
Breath Volume	3.433	08:36
Air Blank	0.000	08:37
Control Test	0.078	08:37
Air Blank	0.000	08:38
Diagnostics Check	OK	08:38

Data Download Date/Time: 07/18/2019 15:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MOONDA,SHUAIB
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 06/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: CHISM

DATE/TIME OF TEST: 07/13/2019 22:02:26

Diagnostics Check	OK	21:51	
Air Blank	0.000	21:51	
Control Test	0.079	21:51	
Air Blank	0.000	21:52	
Subject Sample	# 1	21:55	No Sample Provided
Breath Volume	0.000	21:55	
Air Blank	0.000	21:55	
Air Blank	0.000	21:57	
Subject Sample	# 2	22:01	No Sample Provided
Breath Volume	0.000	22:01	
Air Blank	0.000	22:01	
Control Test	0.079	22:01	
Air Blank	0.000	22:02	
Diagnostics Check	OK	22:02	

Data Download Date/Time: 07/18/2019 15:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MOONDA,SHUIB
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 06/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: LOPEZESCOBAR

DATE/TIME OF TEST: 07/14/2019 01:47:40

Diagnostics Check	OK	01:40
Air Blank	0.000	01:41
Control Test	0.080	01:41
Air Blank	0.000	01:42
Subject Sample # 1	0.166	01:43
Breath Volume	2.886	01:43
Air Blank	0.000	01:44
Air Blank	0.000	01:45
Subject Sample # 2	0.146	01:46
Breath Volume	2.320	01:46
Air Blank	0.000	01:46
Control Test	0.078	01:47
Air Blank	0.000	01:47
Diagnostics Check	OK	01:47

Data Download Date/Time: 07/18/2019 15:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AZEVEDO,FABIO
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 06/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: SWILLEY

DATE/TIME OF TEST: 07/14/2019 02:06:30

Diagnostics Check		OK	01:54	
Air Blank		0.000	01:55	
Control Test		0.079	01:55	
Air Blank		0.000	01:56	
Subject Sample	# 1	0.226	01:59	Volume Not Met
Breath Volume		1.070	01:59	
Air Blank		0.000	02:00	
Air Blank		0.000	02:01	
Subject Sample	# 2	0.189	02:04	Volume Not Met
Breath Volume		0.750	02:04	
Air Blank		0.000	02:05	
Control Test		0.077	02:05	
Air Blank		0.000	02:06	
Diagnostics Check		OK	02:06	

Data Download Date/Time: 07/18/2019 15:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JACKSON, MELISSA
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 06/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: TOOLE

DATE/TIME OF TEST: 07/14/2019 02:58:19

Diagnostics Check		OK	02:48
Air Blank		0.000	02:48
Control Test		0.078	02:48
Air Blank		0.000	02:49
Subject Sample	# 1	0.200	02:50
Breath Volume		1.203	02:50
Air Blank		0.000	02:51
Air Blank		0.000	02:53
Subject Sample	# 2	0.188	02:56
Breath Volume		1.730	02:56
Air Blank		0.000	02:57
Control Test		0.079	02:57
Air Blank		0.000	02:58
Diagnostics Check		OK	02:58

Data Download Date/Time: 07/18/2019 15:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SZESNAT,KENNETH
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 06/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: SANTIAGOJUSINO

DATE/TIME OF TEST: 07/14/2019 05:28:06

Diagnostics Check	OK	05:21
Air Blank	0.000	05:21
Control Test	0.079	05:21
Air Blank	0.000	05:22
Subject Sample # 1	0.138	05:23
Breath Volume	2.164	05:23
Air Blank	0.000	05:23
Air Blank	0.000	05:25
Subject Sample # 2	0.135	05:26
Breath Volume	2.777	05:26
Air Blank	0.000	05:27
Control Test	0.079	05:27
Air Blank	0.000	05:28
Diagnostics Check	OK	05:28

Data Download Date/Time: 07/18/2019 15:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RODRIGUEZ,STEPHANIE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 06/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: WALLS

DATE/TIME OF TEST: 07/15/2019 00:33:19

Diagnostics Check		OK	00:18	
Air Blank		0.000	00:19	
Control Test		0.079	00:19	
Air Blank		0.000	00:20	
Subject Sample	# 1	0.151	00:23	Volume Not Met
Breath Volume		0.046	00:23	
Air Blank		0.000	00:23	
Air Blank		0.000	00:25	
Subject Sample	# 2	0.181	00:27	
Breath Volume		1.699	00:27	
Air Blank		0.000	00:27	
Air Blank		0.000	00:29	
Subject Sample	# 3	0.175	00:31	
Breath Volume		1.707	00:31	
Air Blank		0.000	00:32	
Control Test		0.077	00:32	
Air Blank		0.000	00:33	
Diagnostics Check		OK	00:33	

Data Download Date/Time: 07/18/2019 15:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JORGE, JUNIOR
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 06/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: WESLEY JR

DATE/TIME OF TEST: 07/15/2019 02:21:07

Diagnostics Check	OK	02:17	
Air Blank	0.000	02:18	
Control Test	0.079	02:18	
Air Blank	0.000	02:19	
Subject Sample	# 1	02:19	Subject Test Refused
Breath Volume	0.000	02:19	
Air Blank	0.000	02:20	
Control Test	0.079	02:20	
Air Blank	0.000	02:21	
Diagnostics Check	OK	02:21	

Data Download Date/Time: 07/18/2019 15:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GRAMAZIO, NICHOLAS
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 06/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: JAMES

DATE/TIME OF TEST: 07/15/2019 04:07:52

Diagnostics Check	OK	04:01
Air Blank	0.000	04:02
Control Test	0.079	04:02
Air Blank	0.000	04:02
Subject Sample # 1	0.213	04:03
Breath Volume	2.882	04:03
Air Blank	0.000	04:04
Air Blank	0.000	04:05
Subject Sample # 2	0.198	04:06
Breath Volume	1.617	04:06
Air Blank	0.000	04:06
Control Test	0.078	04:07
Air Blank	0.000	04:07
Diagnostics Check	OK	04:07

Data Download Date/Time: 07/18/2019 15:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PADRON,MARCIE,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LOPEZ,JUSTINA
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 06/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 11:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:28

SUBJECT'S LAST NAME: STUBBS

DATE/TIME OF TEST: 07/15/2019 13:05:18

Diagnostics Check	OK	12:55
Air Blank	0.000	12:56
Control Test	0.080	12:56
Air Blank	0.000	12:56
Subject Sample # 1	0.248	12:58
Breath Volume	1.308	12:58
Air Blank	0.000	12:59
Air Blank	0.000	13:01
Subject Sample # 2	0.238	13:03
Breath Volume	1.175	13:03
Air Blank	0.000	13:04
Control Test	0.078	13:04
Air Blank	0.000	13:05
Diagnostics Check	OK	13:05

Data Download Date/Time: 07/18/2019 15:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BINET,MARLON
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 06/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:51

SUBJECT'S LAST NAME: DIAZ RIOS

DATE/TIME OF TEST: 07/15/2019 21:22:28

Diagnostics Check	OK	21:16
Air Blank	0.000	21:17
Control Test	0.079	21:17
Air Blank	0.000	21:17
Subject Sample # 1	0.000	21:18
Breath Volume	2.042	21:18
Air Blank	0.000	21:18
Air Blank	0.000	21:20
Subject Sample # 2	0.000	21:21
Breath Volume	2.308	21:21
Air Blank	0.000	21:21
Control Test	0.079	21:21
Air Blank	0.000	21:22
Diagnostics Check	OK	21:22

Data Download Date/Time: 07/18/2019 15:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: NELSON, GREGG
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 06/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: SANDOVAL

DATE/TIME OF TEST: 07/16/2019 02:28:47

Diagnostics Check		OK	02:21
Air Blank		0.000	02:22
Control Test		0.079	02:22
Air Blank		0.000	02:23
Subject Sample	# 1	0.198	02:24
Breath Volume		1.433	02:24
Air Blank		0.000	02:24
Air Blank		0.000	02:26
Subject Sample	# 2	0.201	02:27
Breath Volume		1.585	02:27
Air Blank		0.000	02:27
Control Test		0.078	02:28
Air Blank		0.000	02:28
Diagnostics Check		OK	02:28

Data Download Date/Time: 07/18/2019 15:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BINET,MARLON
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 06/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: QUILLEPINEDA

DATE/TIME OF TEST: 07/16/2019 23:27:34

Diagnostics Check	OK	23:18
Air Blank	0.000	23:18
Control Test	0.079	23:19
Air Blank	0.000	23:19
Subject Sample # 1	0.150	23:21
Breath Volume	2.058	23:21
Air Blank	0.000	23:21
Air Blank	0.000	23:23
Subject Sample # 2	0.148	23:25
Breath Volume	2.425	23:25
Air Blank	0.000	23:26
Control Test	0.078	23:26
Air Blank	0.000	23:27
Diagnostics Check	OK	23:27

Data Download Date/Time: 07/18/2019 15:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELVILLE,KELLY,M
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AUSTIN,ADAM
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 06/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 10:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:50

SUBJECT'S LAST NAME: BECKER

DATE/TIME OF TEST: 07/17/2019 12:30:39

Diagnostics Check		OK	12:18	
Air Blank		0.000	12:19	
Control Test		0.080	12:19	
Air Blank		0.000	12:20	
Subject Sample	# 1	0.278	12:23	Volume Not Met
Breath Volume		0.253	12:23	
Air Blank		0.000	12:24	
Air Blank		0.000	12:25	
Subject Sample	# 2	0.251	12:28	Volume Not Met
Breath Volume		0.859	12:28	
Air Blank		0.000	12:29	
Control Test		0.077	12:29	
Air Blank		0.000	12:30	
Diagnostics Check		OK	12:30	

Data Download Date/Time: 07/18/2019 15:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELVILLE,KELLY,M
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AUSTIN,ADAM
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 06/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 10:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:50

SUBJECT'S LAST NAME: BECKER

DATE/TIME OF TEST: 07/17/2019 12:38:48

Diagnostics Check	OK	12:32
Air Blank	0.000	12:32
Control Test	0.076	12:32
Air Blank	0.000	12:33
Subject Sample # 1	0.297	12:33
Breath Volume	1.574	12:33
Air Blank	0.000	12:34
Air Blank	0.000	12:36
Subject Sample # 2	0.284	12:37
Breath Volume	1.246	12:37
Air Blank	0.000	12:37
Control Test	0.077	12:38
Air Blank	0.000	12:38
Diagnostics Check	OK	12:38

Data Download Date/Time: 07/18/2019 15:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WEBB,DENNIS
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 06/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: WEBB

DATE/TIME OF TEST: 07/18/2019 00:18:53

Diagnostics Check	OK	00:16	
Air Blank	0.000	00:16	
Control Test	0.080	00:16	
Air Blank	0.000	00:17	
Subject Sample	# 1	00:17	Subject Test Refused
Breath Volume	0.000	00:17	
Air Blank	0.000	00:17	
Control Test	0.080	00:18	
Air Blank	0.000	00:18	
Diagnostics Check	OK	00:18	

Data Download Date/Time: 07/18/2019 15:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VINER,DENNIS
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 06/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: HAYMAN

DATE/TIME OF TEST: 07/18/2019 03:08:36

Diagnostics Check		OK	03:02
Air Blank		0.000	03:02
Control Test		0.080	03:03
Air Blank		0.000	03:03
Subject Sample	# 1	0.000	03:04
Breath Volume		4.800	03:04
Air Blank		0.000	03:04
Air Blank		0.000	03:06
Subject Sample	# 2	0.000	03:07
Breath Volume		4.996	03:07
Air Blank		0.000	03:07
Control Test		0.079	03:08
Air Blank		0.000	03:08
Diagnostics Check		OK	03:08

Data Download Date/Time: 07/18/2019 15:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HARWOOD,ARMANDO
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 06/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: DAVENPORT

DATE/TIME OF TEST: 07/18/2019 03:50:43

Diagnostics Check	OK	03:44
Air Blank	0.000	03:45
Control Test	0.079	03:45
Air Blank	0.000	03:45
Subject Sample # 1	0.065	03:46
Breath Volume	1.812	03:46
Air Blank	0.000	03:46
Air Blank	0.000	03:48
Subject Sample # 2	0.060	03:49
Breath Volume	1.265	03:49
Air Blank	0.000	03:49
Control Test	0.078	03:50
Air Blank	0.000	03:50
Diagnostics Check	OK	03:50

Data Download Date/Time: 07/18/2019 15:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DONAKOWSKI,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: ALSALHANI

DATE/TIME OF TEST: 07/18/2019 04:22:38

Diagnostics Check	OK	04:19	
Air Blank	0.000	04:19	
Control Test	0.080	04:19	
Air Blank	0.000	04:20	
Subject Sample	# 1	04:21	Subject Test Refused
Breath Volume	0.000	04:21	
Air Blank	0.000	04:21	
Control Test	0.079	04:22	
Air Blank	0.000	04:22	
Diagnostics Check	OK	04:22	

Data Download Date/Time: 07/18/2019 15:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TUERO,ADRIAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: PAULEMAN

DATE/TIME OF TEST: 07/19/2019 04:30:31

Diagnostics Check	OK	04:24
Air Blank	0.000	04:24
Control Test	0.079	04:25
Air Blank	0.000	04:25
Subject Sample # 1	0.234	04:26
Breath Volume	1.214	04:26
Air Blank	0.000	04:26
Air Blank	0.000	04:28
Subject Sample # 2	0.248	04:28
Breath Volume	1.800	04:28
Air Blank	0.000	04:29
Control Test	0.079	04:29
Air Blank	0.000	04:30
Diagnostics Check	OK	04:30

Data Download Date/Time: 07/24/2019 10:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: NORMAN,BJORN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 06/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: ANAND

DATE/TIME OF TEST: 07/20/2019 04:35:35

Diagnostics Check	OK	04:25
Air Blank	0.000	04:26
Control Test	0.080	04:26
Air Blank	0.000	04:27
Subject Sample # 1	0.121	04:28
Breath Volume	1.921	04:28
Air Blank	0.000	04:29
Air Blank	0.000	04:31
Subject Sample # 2	0.115	04:34
Breath Volume	1.988	04:34
Air Blank	0.000	04:34
Control Test	0.078	04:34
Air Blank	0.000	04:35
Diagnostics Check	OK	04:35

Data Download Date/Time: 07/24/2019 10:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GONZALEZ,ASHLEY
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 06/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: SUKMANGAL

DATE/TIME OF TEST: 07/20/2019 23:56:29

Diagnostics Check	OK	23:50
Air Blank	0.000	23:50
Control Test	0.080	23:51
Air Blank	0.000	23:51
Subject Sample # 1	0.169	23:52
Breath Volume	1.832	23:52
Air Blank	0.000	23:52
Air Blank	0.000	23:54
Subject Sample # 2	0.169	23:54
Breath Volume	1.753	23:54
Air Blank	0.000	23:55
Control Test	0.079	23:55
Air Blank	0.000	23:56
Diagnostics Check	OK	23:56

Data Download Date/Time: 07/24/2019 10:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MONTGOMERY, JAMES
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 06/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: FABER

DATE/TIME OF TEST: 07/21/2019 02:50:00

Diagnostics Check	OK	02:43
Air Blank	0.000	02:44
Control Test	0.080	02:44
Air Blank	0.000	02:44
Subject Sample # 1	0.107	02:45
Breath Volume	3.511	02:45
Air Blank	0.000	02:46
Air Blank	0.000	02:48
Subject Sample # 2	0.103	02:48
Breath Volume	2.300	02:48
Air Blank	0.000	02:49
Control Test	0.078	02:49
Air Blank	0.000	02:49
Diagnostics Check	OK	02:50

Data Download Date/Time: 07/24/2019 10:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: THOMAS,RICK
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 06/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: LEDFORD

DATE/TIME OF TEST: 07/21/2019 03:42:57

Diagnostics Check		OK	03:36
Air Blank		0.000	03:37
Control Test		0.079	03:37
Air Blank		0.000	03:38
Subject Sample	# 1	0.198	03:38
Breath Volume		2.750	03:38
Air Blank		0.000	03:39
Air Blank		0.000	03:40
Subject Sample	# 2	0.196	03:41
Breath Volume		2.859	03:41
Air Blank		0.000	03:42
Control Test		0.078	03:42
Air Blank		0.000	03:42
Diagnostics Check		OK	03:42

Data Download Date/Time: 07/24/2019 10:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BINET, MARLON
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 06/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: LEAL CARRILLO

DATE/TIME OF TEST: 07/21/2019 04:10:31

Diagnostics Check	OK	04:03
Air Blank	0.000	04:04
Control Test	0.079	04:04
Air Blank	0.000	04:05
Subject Sample # 1	0.113	04:06
Breath Volume	2.109	04:06
Air Blank	0.000	04:06
Air Blank	0.000	04:08
Subject Sample # 2	0.113	04:09
Breath Volume	1.765	04:09
Air Blank	0.000	04:09
Control Test	0.078	04:09
Air Blank	0.000	04:10
Diagnostics Check	OK	04:10

Data Download Date/Time: 07/24/2019 10:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AMARAL,WILKER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 07/21/2019 05:15:09

Diagnostics Check	OK	05:08
Air Blank	0.000	05:08
Control Test	0.080	05:08
Air Blank	0.000	05:09
Subject Sample # 1	0.291	05:10
Breath Volume	1.531	05:10
Air Blank	0.000	05:11
Air Blank	0.000	05:12
Subject Sample # 2	0.276	05:13
Breath Volume	1.792	05:13
Air Blank	0.000	05:14
Control Test	0.078	05:14
Air Blank	0.000	05:15
Diagnostics Check	OK	05:15

Data Download Date/Time: 07/24/2019 10:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HARMAN,RICHARD
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: DOMINGO BRAVO

DATE/TIME OF TEST: 07/21/2019 23:58:00

Diagnostics Check	OK	23:51
Air Blank	0.000	23:52
Control Test	0.079	23:52
Air Blank	0.000	23:53
Subject Sample # 1	0.159	23:53
Breath Volume	1.917	23:53
Air Blank	0.000	23:54
Air Blank	0.000	23:56
Subject Sample # 2	0.157	23:56
Breath Volume	1.750	23:56
Air Blank	0.000	23:57
Control Test	0.078	23:57
Air Blank	0.000	23:57
Diagnostics Check	OK	23:58

Data Download Date/Time: 07/24/2019 10:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: FIGUEROA,HECTOR
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 06/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: SANTIAGOSANTIAGO

DATE/TIME OF TEST: 07/22/2019 03:38:32

Diagnostics Check	OK	03:31
Air Blank	0.000	03:32
Control Test	0.080	03:32
Air Blank	0.000	03:33
Subject Sample # 1	0.082	03:34
Breath Volume	1.316	03:34
Air Blank	0.000	03:34
Air Blank	0.000	03:36
Subject Sample # 2	0.083	03:37
Breath Volume	3.441	03:37
Air Blank	0.000	03:37
Control Test	0.079	03:37
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

Data Download Date/Time: 07/24/2019 10:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MENDOZA,ANGEL
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 06/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: GOMEZGOMEZ

DATE/TIME OF TEST: 07/24/2019 01:02:16

Diagnostics Check	OK	00:56
Air Blank	0.000	00:56
Control Test	0.080	00:57
Air Blank	0.000	00:57
Subject Sample # 1	0.000	00:58
Breath Volume	1.265	00:58
Air Blank	0.000	00:58
Air Blank	0.000	01:00
Subject Sample # 2	0.000	01:00
Breath Volume	2.335	01:00
Air Blank	0.000	01:01
Control Test	0.080	01:01
Air Blank	0.000	01:02
Diagnostics Check	OK	01:02

Data Download Date/Time: 07/24/2019 10:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

INSTRUMENT REGISTERED TO: COCOA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,JAMES,H
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: JOHNSON,K
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 06/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 07/19/2019 23:15:06

Diagnostics Check	OK	23:12	
Air Blank	0.000	23:12	
Control Test	0.078	23:13	
Air Blank	0.000	23:13	
Subject Sample	# 1	23:13	Subject Test Refused
Breath Volume	0.000	23:13	
Air Blank	0.000	23:14	
Control Test	0.078	23:14	
Air Blank	0.000	23:15	
Diagnostics Check	OK	23:15	

Data Download Date/Time: 07/30/2019 21:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

INSTRUMENT REGISTERED TO: COCOA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANKS,RICKESHIA,T
BREATH TEST OPERATOR AGENCY: Rockledge PD
ARRESTING OFFICER: SANCHEZ-SIERRA,NATALIA
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 06/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: MAITA

DATE/TIME OF TEST: 07/21/2019 01:54:57

Diagnostics Check	OK	01:53	
Air Blank	0.000	01:54	
Control Test	0.074	01:54	Control Outside Tolerance
Air Blank	0.000	01:54	

Data Download Date/Time: 07/30/2019 21:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

INSTRUMENT REGISTERED TO: COCOA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANKS,RICKESHIA,T
BREATH TEST OPERATOR AGENCY: Rockledge PD
ARRESTING OFFICER: SANCHEZ-SIERRA,NATALIA
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 06/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: MIATA

DATE/TIME OF TEST: 07/21/2019 02:00:02

Diagnostics Check	OK	01:57	
Air Blank	0.000	01:57	
Control Test	0.075	01:57	
Air Blank	0.000	01:58	
Subject Sample	# 1	01:58	Subject Test Refused
Breath Volume	0.000	01:58	
Air Blank	0.000	01:59	
Control Test	0.075	01:59	
Air Blank	0.000	01:59	
Diagnostics Check	OK	02:00	

Data Download Date/Time: 07/30/2019 21:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

INSTRUMENT REGISTERED TO: COCOA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAVALIERI, MICHAEL, A
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: PENDER, DERICK
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 06/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 07:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:50

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 07/27/2019 08:32:29

Diagnostics Check	OK	08:25
Air Blank	0.000	08:25
Control Test	0.078	08:26
Air Blank	0.000	08:26
Subject Sample # 1	0.057	08:27
Breath Volume	2.304	08:27
Air Blank	0.000	08:28
Air Blank	0.000	08:30
Subject Sample # 2	0.052	08:30
Breath Volume	1.503	08:30
Air Blank	0.000	08:31
Control Test	0.078	08:31
Air Blank	0.000	08:32
Diagnostics Check	OK	08:32

Data Download Date/Time: 07/30/2019 21:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

INSTRUMENT REGISTERED TO: COCOA P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HANSON,JASON,A

BREATH TEST OPERATOR AGENCY: Titusville PD

ARRESTING OFFICER: BOND,BRAIN

ARRESTING OFFICER AGENCY: Titusville PD

LAST AGENCY INSPECTION: 06/25/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 02618080A3

DRY GAS EXPIRY DATE: 03/05/2020

VIOLATION CODE: DUI

ARREST TIME: 22:11

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:46

SUBJECT'S LAST NAME: FUERTES

DATE/TIME OF TEST: 07/28/2019 23:32:21

Diagnostics Check	OK	23:22
Air Blank	0.000	23:23
Control Test	0.078	23:23
Air Blank	0.000	23:24
Subject Sample # 1	0.073	23:26
Breath Volume	1.500	23:26
Air Blank	0.000	23:26
Air Blank	0.000	23:28
Subject Sample # 2	0.075	23:30
Breath Volume	1.863	23:30
Air Blank	0.000	23:31
Control Test	0.076	23:31
Air Blank	0.000	23:32
Diagnostics Check	OK	23:32

Data Download Date/Time: 07/30/2019 21:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD, WILLIAM,
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: HOWARD, WILLIAM
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: GETZ

DATE/TIME OF TEST: 07/05/2019 01:47:19

Diagnostics Check	OK	01:40
Air Blank	0.000	01:41
Control Test	0.079	01:41
Air Blank	0.000	01:42
Subject Sample # 1	0.121	01:42
Breath Volume	1.781	01:42
Air Blank	0.000	01:43
Air Blank	0.000	01:45
Subject Sample # 2	0.122	01:45
Breath Volume	1.574	01:45
Air Blank	0.000	01:46
Control Test	0.080	01:46
Air Blank	0.000	01:47
Diagnostics Check	OK	01:47

Data Download Date/Time: 07/30/2019 07:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,ANTHONY,M
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: TARANTINO,G
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 06/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 07/21/2019 04:19:10

Diagnostics Check	OK	04:11
Air Blank	0.000	04:11
Control Test	0.080	04:12
Air Blank	0.000	04:12
Subject Sample # 1	0.135	04:14
Breath Volume	2.925	04:14
Air Blank	0.000	04:14
Air Blank	0.000	04:16
Subject Sample # 2	0.125	04:17
Breath Volume	2.367	04:17
Air Blank	0.000	04:18
Control Test	0.079	04:18
Air Blank	0.000	04:19
Diagnostics Check	OK	04:19

Data Download Date/Time: 07/30/2019 07:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOUBEK,STEPHEN,
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: AIANI,JOSEPH
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 06/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: RICKETT

DATE/TIME OF TEST: 07/28/2019 01:18:17

Diagnostics Check	OK	01:10
Air Blank	0.000	01:11
Control Test	0.080	01:11
Air Blank	0.000	01:12
Subject Sample # 1	0.063	01:13
Breath Volume	1.753	01:13
Air Blank	0.000	01:14
Air Blank	0.000	01:16
Subject Sample # 2	0.063	01:16
Breath Volume	3.121	01:16
Air Blank	0.000	01:17
Control Test	0.079	01:17
Air Blank	0.000	01:18
Diagnostics Check	OK	01:18

Data Download Date/Time: 07/30/2019 07:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOUBEK,STEPHEN,
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: IERVASI,MAURIZIO
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 07/29/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 20:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: NORMAN

DATE/TIME OF TEST: 07/31/2019 21:34:51

Diagnostics Check	OK	21:23
Air Blank	0.000	21:23
Control Test	0.081	21:23
Air Blank	0.000	21:24
Subject Sample # 1	0.258	21:27
Breath Volume	1.875	21:27
Air Blank	0.000	21:28
Air Blank	0.000	21:29
Subject Sample # 2	0.272	21:33
Breath Volume	1.914	21:33
Air Blank	0.000	21:33
Control Test	0.077	21:34
Air Blank	0.000	21:34
Diagnostics Check	OK	21:34

Data Download Date/Time: 08/05/2019 16:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

INSTRUMENT REGISTERED TO: INDIAN HARBOUR PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,JIM,E
BREATH TEST OPERATOR AGENCY: Indian Harbour Beach
ARRESTING OFFICER: AIANI,JOSEPH
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 06/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: BELCHER

DATE/TIME OF TEST: 07/04/2019 06:04:51

Diagnostics Check	OK	05:58
Air Blank	0.000	05:59
Control Test	0.080	05:59
Air Blank	0.000	06:00
Subject Sample # 1	0.000	06:00
Breath Volume	3.636	06:00
Air Blank	0.000	06:01
Air Blank	0.000	06:03
Subject Sample # 2	0.000	06:03
Breath Volume	3.828	06:03
Air Blank	0.000	06:03
Control Test	0.079	06:04
Air Blank	0.000	06:04
Diagnostics Check	OK	06:04

Data Download Date/Time: 07/26/2019 19:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

INSTRUMENT REGISTERED TO: INDIAN HARBOUR PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGRORY,DIARMUID,J
BREATH TEST OPERATOR AGENCY: Indian Harbour Beach
ARRESTING OFFICER: COCCHIOLA,AUSTIN
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 06/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 07/12/2019 03:32:41

Diagnostics Check	OK	03:25
Air Blank	0.000	03:26
Control Test	0.080	03:26
Air Blank	0.000	03:26
Subject Sample # 1	0.189	03:27
Breath Volume	2.511	03:27
Air Blank	0.000	03:28
Air Blank	0.000	03:30
Subject Sample # 2	0.176	03:31
Breath Volume	1.289	03:31
Air Blank	0.000	03:31
Control Test	0.080	03:32
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

Data Download Date/Time: 07/26/2019 19:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

INSTRUMENT REGISTERED TO: INDIAN HARBOUR PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGRORY,DIARMUID,J
BREATH TEST OPERATOR AGENCY: Indian Harbour Beach
ARRESTING OFFICER: PINSKER,JOSEPH
ARRESTING OFFICER AGENCY: Indian Harbour Beach
LAST AGENCY INSPECTION: 06/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: LANGE

DATE/TIME OF TEST: 07/26/2019 02:41:01

Diagnostics Check	OK	02:33
Air Blank	0.000	02:34
Control Test	0.080	02:34
Air Blank	0.000	02:34
Subject Sample # 1	0.222	02:36
Breath Volume	1.500	02:36
Air Blank	0.000	02:36
Air Blank	0.000	02:38
Subject Sample # 2	0.217	02:39
Breath Volume	2.195	02:39
Air Blank	0.000	02:40
Control Test	0.079	02:40
Air Blank	0.000	02:40
Diagnostics Check	OK	02:41

Data Download Date/Time: 07/26/2019 19:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

INSTRUMENT REGISTERED TO: INDIAN HARBOUR PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD, WILLIM,
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: HOWARD, WILLIAM
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 07/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 07/27/2019 23:27:05

Diagnostics Check	OK	23:20
Air Blank	0.000	23:20
Control Test	0.080	23:20
Air Blank	0.000	23:21
Subject Sample # 1	0.167	23:22
Breath Volume	1.210	23:22
Air Blank	0.000	23:22
Air Blank	0.000	23:24
Subject Sample # 2	0.158	23:25
Breath Volume	1.425	23:25
Air Blank	0.000	23:26
Control Test	0.080	23:26
Air Blank	0.000	23:27
Diagnostics Check	OK	23:27

Data Download Date/Time: 08/25/2019 04:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIVINS, HEATHER, C
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: LIGHTSEY, DUSTIN
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 06/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A3
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: BUI
ARREST TIME: 17:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:43

SUBJECT'S LAST NAME: MCCOLLOUGH

DATE/TIME OF TEST: 07/06/2019 18:31:42

Diagnostics Check	OK	18:25
Air Blank	0.000	18:25
Control Test	0.080	18:26
Air Blank	0.000	18:26
Subject Sample # 1	0.139	18:26
Breath Volume	2.421	18:26
Air Blank	0.000	18:27
Air Blank	0.000	18:29
Subject Sample # 2	0.135	18:30
Breath Volume	1.863	18:30
Air Blank	0.000	18:30
Control Test	0.078	18:31
Air Blank	0.000	18:31
Diagnostics Check	OK	18:31

Data Download Date/Time: 07/28/2019 22:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINSTEAD,JUSTIN,M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: ITRI,ANGELA
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A3
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: MORRELL

DATE/TIME OF TEST: 07/07/2019 20:49:18

Diagnostics Check	OK	20:41
Air Blank	0.000	20:41
Control Test	0.082	20:41
Air Blank	0.000	20:42
Subject Sample # 1	0.061	20:44
Breath Volume	1.937	20:44
Air Blank	0.000	20:45
Air Blank	0.000	20:47
Subject Sample # 2	0.062	20:47
Breath Volume	2.132	20:47
Air Blank	0.000	20:48
Control Test	0.081	20:48
Air Blank	0.000	20:49
Diagnostics Check	OK	20:49

Data Download Date/Time: 07/28/2019 22:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINSTEAD,JUSTIN,M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: GRAY,CONNOR
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A3
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:47

SUBJECT'S LAST NAME: MOSER

DATE/TIME OF TEST: 07/07/2019 22:33:22

Diagnostics Check	OK	22:25
Air Blank	0.000	22:26
Control Test	0.080	22:26
Air Blank	0.000	22:27
Subject Sample # 1	0.330	22:28
Breath Volume	2.843	22:28
Air Blank	0.000	22:29
Air Blank	0.000	22:30
Subject Sample # 2	0.328	22:31
Breath Volume	2.718	22:31
Air Blank	0.000	22:32
Control Test	0.080	22:32
Air Blank	0.000	22:33
Diagnostics Check	OK	22:33

Data Download Date/Time: 07/28/2019 22:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,NICHOLAS,W
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: ROSA,E
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 06/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A3
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: SOMERS

DATE/TIME OF TEST: 07/09/2019 22:36:15

Diagnostics Check	OK	22:30
Air Blank	0.000	22:30
Control Test	0.081	22:30
Air Blank	0.000	22:31
Subject Sample # 1	0.095	22:31
Breath Volume	1.933	22:31
Air Blank	0.000	22:32
Air Blank	0.000	22:34
Subject Sample # 2	0.099	22:34
Breath Volume	1.324	22:34
Air Blank	0.000	22:35
Control Test	0.081	22:35
Air Blank	0.000	22:36
Diagnostics Check	OK	22:36

Data Download Date/Time: 07/28/2019 22:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINSTEAD,JUSTIN,M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: WINSTEAD,JUSTIN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A3
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: LOCKLEAR

DATE/TIME OF TEST: 07/11/2019 01:12:31

Diagnostics Check		OK	01:05
Air Blank		0.000	01:06
Control Test		0.081	01:06
Air Blank		0.000	01:06
Subject Sample	# 1	0.204	01:08
Breath Volume		2.273	01:08
Air Blank		0.000	01:08
Air Blank		0.000	01:10
Subject Sample	# 2	0.208	01:11
Breath Volume		1.449	01:11
Air Blank		0.000	01:11
Control Test		0.080	01:11
Air Blank		0.000	01:12
Diagnostics Check		OK	01:12

Data Download Date/Time: 07/28/2019 22:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,ROBERT,F
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: WAGNER,ROBERT
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A3
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: TANNER

DATE/TIME OF TEST: 07/13/2019 03:20:46

Diagnostics Check	OK	03:13
Air Blank	0.000	03:14
Control Test	0.081	03:14
Air Blank	0.000	03:15
Subject Sample # 1	0.131	03:15
Breath Volume	1.121	03:15
Air Blank	0.000	03:16
Air Blank	0.000	03:18
Subject Sample # 2	0.140	03:19
Breath Volume	1.390	03:19
Air Blank	0.000	03:19
Control Test	0.080	03:20
Air Blank	0.000	03:20
Diagnostics Check	OK	03:20

Data Download Date/Time: 07/28/2019 22:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINSTEAD,JUSTIN,M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: ITRI,ANGELA
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A3
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 07/16/2019 00:20:28

Diagnostics Check	OK	00:17	
Air Blank	0.000	00:18	
Control Test	0.082	00:18	
Air Blank	0.000	00:18	
Subject Sample	# 1	00:19	Subject Test Refused
Breath Volume	0.000	00:19	
Air Blank	0.000	00:19	
Control Test	0.082	00:19	
Air Blank	0.000	00:20	
Diagnostics Check	OK	00:20	

Data Download Date/Time: 07/28/2019 22:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBIN,GASTON,A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: QUIMBAYO,MONICA
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A3
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 07/19/2019 04:56:55

Diagnostics Check		OK	04:50
Air Blank		0.000	04:50
Control Test		0.080	04:51
Air Blank		0.000	04:51
Subject Sample	# 1	0.156	04:52
Breath Volume		2.617	04:52
Air Blank		0.000	04:52
Air Blank		0.000	04:54
Subject Sample	# 2	0.157	04:55
Breath Volume		2.703	04:55
Air Blank		0.000	04:56
Control Test		0.080	04:56
Air Blank		0.000	04:56
Diagnostics Check		OK	04:56

Data Download Date/Time: 07/28/2019 22:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,ROBERT,F
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: WAGNER,ROBERT
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A3
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 07/21/2019 01:28:19

Diagnostics Check	OK	01:21
Air Blank	0.000	01:22
Control Test	0.081	01:22
Air Blank	0.000	01:23
Subject Sample # 1	0.185	01:23
Breath Volume	1.203	01:23
Air Blank	0.000	01:24
Air Blank	0.000	01:26
Subject Sample # 2	0.186	01:26
Breath Volume	1.460	01:26
Air Blank	0.000	01:27
Control Test	0.080	01:27
Air Blank	0.000	01:28
Diagnostics Check	OK	01:28

Data Download Date/Time: 07/28/2019 22:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBIN,GASTON,A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: THOMAN,TYLER
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A3
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: CHRISTIAN

DATE/TIME OF TEST: 07/27/2019 03:57:17

Diagnostics Check	OK	03:49
Air Blank	0.000	03:50
Control Test	0.080	03:50
Air Blank	0.000	03:51
Subject Sample # 1	0.126	03:52
Breath Volume	1.789	03:52
Air Blank	0.000	03:53
Air Blank	0.000	03:54
Subject Sample # 2	0.133	03:55
Breath Volume	2.015	03:55
Air Blank	0.000	03:56
Control Test	0.080	03:56
Air Blank	0.000	03:57
Diagnostics Check	OK	03:57

Data Download Date/Time: 07/28/2019 22:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JENKINS,CHRISTINA,E
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: JENKINS,CHRISTINA
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 07/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A3
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: LEWIS SR

DATE/TIME OF TEST: 07/29/2019 22:29:13

Diagnostics Check		OK	22:17	
Air Blank		0.000	22:18	
Control Test		0.081	22:18	
Air Blank		0.000	22:18	
Subject Sample	# 1	0.308	22:22	Volume Not Met
Breath Volume		0.019	22:22	
Air Blank		0.000	22:22	
Air Blank		0.000	22:24	
Subject Sample	# 2	0.266	22:27	Volume Not Met
Breath Volume		0.015	22:27	
Air Blank		0.000	22:28	
Control Test		0.073	22:28	Control Outside Tolerance
Air Blank		0.000	22:29	

Data Download Date/Time: 08/19/2019 21:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JENKINS,CHRISTINA,E
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: APPEGATE,ALEXANDRIA
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 07/28/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 19817080A3
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/29/2019 22:57:52

Diagnostics Check	OK	22:56	
Air Blank	0.000	22:56	
Control Test	0.077	22:57	
Air Blank		22:57	RFI Detect
Air Blank	0.000	22:57	

Data Download Date/Time: 08/19/2019 21:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JENKINS,CHRISTINA,E
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: APPLGATE,ALEXANDRIA
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 07/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A3
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/29/2019 23:09:12

Diagnostics Check		OK	23:02
Air Blank		0.000	23:03
Control Test		0.078	23:03
Air Blank		0.000	23:03
Subject Sample	# 1	0.150	23:04
Breath Volume		3.679	23:04
Air Blank		0.000	23:05
Air Blank		0.000	23:07
Subject Sample	# 2	0.150	23:07
Breath Volume		3.695	23:07
Air Blank		0.000	23:08
Control Test		0.077	23:08
Air Blank		0.000	23:09
Diagnostics Check		OK	23:09

Data Download Date/Time: 08/19/2019 21:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINSTEAD,JUSTIN,M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BIALOBRZESKI,MATTHEW
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 07/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A3
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: RASHID

DATE/TIME OF TEST: 07/30/2019 23:05:15

Diagnostics Check	OK	22:54
Air Blank	0.000	22:55
Control Test	0.079	22:55
Air Blank	0.000	22:56
Subject Sample # 1	0.103	22:59
Breath Volume	1.562	22:59
Air Blank	0.000	23:00
Air Blank	0.000	23:01
Subject Sample # 2	0.102	23:03
Breath Volume	1.839	23:03
Air Blank	0.000	23:04
Control Test	0.078	23:04
Air Blank	0.000	23:05
Diagnostics Check	OK	23:05

Data Download Date/Time: 08/19/2019 21:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 06/28/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24818080A2

DRY GAS EXPIRY DATE: 10/05/2020

VIOLATION CODE: DUI

ARREST TIME: 23:23

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: FERRAGAMO

DATE/TIME OF TEST: 07/02/2019 23:45:17

Diagnostics Check		OK	23:30	
Air Blank		0.000	23:31	
Control Test		0.082	23:31	
Air Blank		0.000	23:31	
Subject Sample	# 1	0.320	23:32	
Breath Volume		2.164	23:32	
Air Blank		0.000	23:33	
Air Blank		0.000	23:34	
Subject Sample	# 2	0.216	23:38	Volume Not Met
Breath Volume		0.175	23:38	
Air Blank		0.000	23:38	
Air Blank		0.000	23:40	
Subject Sample	# 3	0.190	23:43	Volume Not Met
Breath Volume		0.363	23:43	
Air Blank		0.000	23:44	
Control Test		0.079	23:44	
Air Blank		0.000	23:45	
Diagnostics Check		OK	23:45	

Data Download Date/Time: 07/29/2019 02:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 24818080A2
BREATH TEST OPERATOR: LUTZ,EDWIN,J	DRY GAS EXPIRY DATE: 10/05/2020
BREATH TEST OPERATOR AGENCY: Palm Bay PD	VIOLATION CODE: DUI
ARRESTING OFFICER: LUTZ,EDWIN	ARREST TIME: 19:04
ARRESTING OFFICER AGENCY: Palm Bay PD	GENDER: Female
LAST AGENCY INSPECTION: 06/28/2019	OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: GONZALEZ TIRADO DATE/TIME OF TEST: 07/06/2019 19:27:03

Diagnostics Check	OK	19:15	
Air Blank	0.000	19:16	
Control Test	0.081	19:16	
Air Blank	0.000	19:17	
Subject Sample	# 1	19:20	No Sample Provided
Breath Volume	0.000	19:20	
Air Blank	0.000	19:20	
Air Blank	0.000	19:22	
Subject Sample	# 2	19:25	Subject Test Refused
Breath Volume	0.000	19:25	
Air Blank	0.000	19:26	
Control Test	0.080	19:26	
Air Blank	0.000	19:26	
Diagnostics Check	OK	19:27	

Data Download Date/Time: 07/29/2019 02:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUTZ,EDWIN,J
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: LUTZ,EDWIN
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 06/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: GONZALEZ TIRADO

DATE/TIME OF TEST: 07/06/2019 19:46:27

Diagnostics Check		OK	19:34	
Air Blank		0.000	19:35	
Control Test		0.079	19:35	
Air Blank		0.000	19:36	
Subject Sample	# 1	0.084	19:39	Volume Not Met
Breath Volume		0.566	19:39	
Air Blank		0.000	19:39	
Air Blank		0.000	19:41	
Subject Sample	# 2	0.080	19:44	Volume Not Met
Breath Volume		0.933	19:44	
Air Blank		0.000	19:45	
Control Test		0.080	19:45	
Air Blank		0.000	19:46	
Diagnostics Check		OK	19:46	

Data Download Date/Time: 07/29/2019 02:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 06/28/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24818080A2

DRY GAS EXPIRY DATE: 10/05/2020

VIOLATION CODE: DUI

ARREST TIME: 04:12

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: MORERA-SANTAMARIA

DATE/TIME OF TEST: 07/07/2019 04:31:07

Diagnostics Check	OK	04:24
Air Blank	0.000	04:25
Control Test	0.078	04:25
Air Blank	0.000	04:25
Subject Sample # 1	0.044	04:26
Breath Volume	1.734	04:26
Air Blank	0.000	04:26
Air Blank	0.000	04:28
Subject Sample # 2	0.046	04:29
Breath Volume	1.968	04:29
Air Blank	0.000	04:30
Control Test	0.080	04:30
Air Blank	0.000	04:31
Diagnostics Check	OK	04:31

Data Download Date/Time: 07/29/2019 02:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUTZ,EDWIN,J
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: LUTZ,EDWIN
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 06/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:58

SUBJECT'S LAST NAME: VELAZQUEZ GARCIA

DATE/TIME OF TEST: 07/07/2019 19:33:53

Diagnostics Check		OK	19:22	
Air Blank		0.000	19:22	
Control Test		0.079	19:23	
Air Blank		0.000	19:23	
Subject Sample	# 1	0.123	19:26	Volume Not Met
Breath Volume		0.019	19:26	
Air Blank		0.000	19:27	
Air Blank		0.000	19:29	
Subject Sample	# 2	0.108	19:32	Volume Not Met
Breath Volume		0.027	19:32	
Air Blank		0.000	19:32	
Control Test		0.079	19:33	
Air Blank		0.000	19:33	
Diagnostics Check		OK	19:33	

Data Download Date/Time: 07/29/2019 02:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUTZ,EDWIN,J
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: MILIAN,S
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 06/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: AGUINAGA

DATE/TIME OF TEST: 07/07/2019 22:40:31

Diagnostics Check	OK	22:34
Air Blank	0.000	22:34
Control Test	0.079	22:35
Air Blank	0.000	22:35
Subject Sample # 1	0.188	22:36
Breath Volume	1.191	22:36
Air Blank	0.000	22:36
Air Blank	0.000	22:38
Subject Sample # 2	0.198	22:39
Breath Volume	1.578	22:39
Air Blank	0.000	22:39
Control Test	0.079	22:39
Air Blank	0.000	22:40
Diagnostics Check	OK	22:40

Data Download Date/Time: 07/29/2019 02:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 06/28/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24818080A2

DRY GAS EXPIRY DATE: 10/05/2020

VIOLATION CODE: DUI

ARREST TIME: 00:42

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 07/14/2019 00:54:09

Diagnostics Check		OK	00:48
Air Blank		0.000	00:48
Control Test		0.080	00:48
Air Blank		0.000	00:49
Subject Sample	# 1	0.182	00:49
Breath Volume		3.015	00:49
Air Blank		0.000	00:50
Air Blank		0.000	00:52
Subject Sample	# 2	0.184	00:52
Breath Volume		2.906	00:52
Air Blank		0.000	00:53
Control Test		0.080	00:53
Air Blank		0.000	00:54
Diagnostics Check		OK	00:54

Data Download Date/Time: 07/29/2019 02:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 06/28/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24818080A2

DRY GAS EXPIRY DATE: 10/05/2020

VIOLATION CODE: DUI

ARREST TIME: 23:31

GENDER: Female

OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: RAWDON

DATE/TIME OF TEST: 07/18/2019 23:47:46

Diagnostics Check	OK	23:40
Air Blank	0.000	23:40
Control Test	0.082	23:41
Air Blank	0.000	23:41
Subject Sample # 1	0.113	23:43
Breath Volume	2.226	23:43
Air Blank	0.000	23:44
Air Blank	0.000	23:45
Subject Sample # 2	0.121	23:46
Breath Volume	2.269	23:46
Air Blank	0.000	23:46
Control Test	0.080	23:47
Air Blank	0.000	23:47
Diagnostics Check	OK	23:47

Data Download Date/Time: 07/29/2019 02:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 06/28/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24818080A2

DRY GAS EXPIRY DATE: 10/05/2020

VIOLATION CODE: DUI

ARREST TIME: 02:31

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: FETTERLY

DATE/TIME OF TEST: 07/19/2019 02:50:17

Diagnostics Check	OK	02:44
Air Blank	0.000	02:44
Control Test	0.080	02:44
Air Blank	0.000	02:45
Subject Sample # 1	0.130	02:45
Breath Volume	1.328	02:45
Air Blank	0.000	02:46
Air Blank	0.000	02:48
Subject Sample # 2	0.136	02:48
Breath Volume	2.140	02:48
Air Blank	0.000	02:49
Control Test	0.079	02:49
Air Blank	0.000	02:50
Diagnostics Check	OK	02:50

Data Download Date/Time: 07/29/2019 02:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 06/28/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24818080A2

DRY GAS EXPIRY DATE: 10/05/2020

VIOLATION CODE: DUI

ARREST TIME: 03:20

GENDER: Female

OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: ACOR

DATE/TIME OF TEST: 07/20/2019 04:34:22

Diagnostics Check		OK	04:22	
Air Blank		0.000	04:23	
Control Test		0.080	04:23	
Air Blank		0.000	04:24	
Subject Sample	# 1	0.165	04:27	Volume Not Met
Breath Volume		0.617	04:27	
Air Blank		0.000	04:27	
Air Blank		0.000	04:29	
Subject Sample	# 2	0.182	04:32	Volume Not Met
Breath Volume		0.406	04:32	
Air Blank		0.000	04:33	
Control Test		0.079	04:33	
Air Blank		0.000	04:34	
Diagnostics Check		OK	04:34	

Data Download Date/Time: 07/29/2019 02:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 06/28/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24818080A2

DRY GAS EXPIRY DATE: 10/05/2020

VIOLATION CODE: DUI

ARREST TIME: 02:21

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: BALTER

DATE/TIME OF TEST: 07/21/2019 02:43:50

Diagnostics Check		OK	02:31
Air Blank		0.000	02:31
Control Test		0.080	02:31
Air Blank		0.000	02:32
Subject Sample	# 1	0.242	02:35
Breath Volume		1.132	02:35
Air Blank		0.000	02:35
Air Blank		0.000	02:37
Subject Sample	# 2	0.216	02:39
Breath Volume		1.207	02:39
Air Blank		0.000	02:39
Air Blank		0.000	02:41
Subject Sample	# 3	0.235	02:42
Breath Volume		1.781	02:42
Air Blank		0.000	02:42
Control Test		0.078	02:43
Air Blank		0.000	02:43
Diagnostics Check		OK	02:43

Data Download Date/Time: 07/29/2019 02:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRENCH,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Sanford PD
ARRESTING OFFICER: PEREZ,JAVIER
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 05:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:00

SUBJECT'S LAST NAME: FAJARDO

DATE/TIME OF TEST: 07/03/2019 07:36:31

Diagnostics Check		OK	07:28
Air Blank		0.000	07:29
Control Test		0.080	07:29
Air Blank		0.000	07:30
Subject Sample	# 1	0.218	07:31
Breath Volume		1.750	07:31
Air Blank		0.000	07:31
Air Blank		0.000	07:33
Subject Sample	# 2	0.212	07:34
Breath Volume		1.355	07:34
Air Blank		0.000	07:35
Control Test		0.080	07:35
Air Blank		0.000	07:36
Diagnostics Check		OK	07:36

Data Download Date/Time: 08/03/2019 04:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA, RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: PRICE, ZACHARY
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: WARGO

DATE/TIME OF TEST: 07/04/2019 02:19:05

Diagnostics Check	OK	02:12
Air Blank	0.000	02:13
Control Test	0.080	02:13
Air Blank	0.000	02:14
Subject Sample # 1	0.298	02:14
Breath Volume	2.343	02:14
Air Blank	0.000	02:15
Air Blank	0.000	02:17
Subject Sample # 2	0.289	02:17
Breath Volume	1.859	02:17
Air Blank	0.000	02:18
Control Test	0.080	02:18
Air Blank	0.000	02:18
Diagnostics Check	OK	02:19

Data Download Date/Time: 08/03/2019 04:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRENCH,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Sanford PD
ARRESTING OFFICER: NUNEZ,JONATHAN
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 05:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:20

SUBJECT'S LAST NAME: CADET

DATE/TIME OF TEST: 07/09/2019 07:01:29

Diagnostics Check	OK	06:54
Air Blank	0.000	06:54
Control Test	0.081	06:55
Air Blank	0.000	06:55
Subject Sample # 1	0.169	06:56
Breath Volume	1.867	06:56
Air Blank	0.000	06:57
Air Blank	0.000	06:59
Subject Sample # 2	0.160	06:59
Breath Volume	1.292	06:59
Air Blank	0.000	07:00
Control Test	0.080	07:00
Air Blank	0.000	07:01
Diagnostics Check	OK	07:01

Data Download Date/Time: 08/03/2019 04:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: CANNADAY,ZACHERY
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 07/12/2019 04:40:35

Diagnostics Check	OK	04:34
Air Blank	0.000	04:34
Control Test	0.081	04:35
Air Blank	0.000	04:35
Subject Sample # 1	0.107	04:36
Breath Volume	2.882	04:36
Air Blank	0.000	04:36
Air Blank	0.000	04:38
Subject Sample # 2	0.102	04:39
Breath Volume	2.390	04:39
Air Blank	0.000	04:39
Control Test	0.080	04:39
Air Blank	0.000	04:40
Diagnostics Check	OK	04:40

Data Download Date/Time: 08/03/2019 04:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA, RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: HOLLAND, TROY
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: DIXON

DATE/TIME OF TEST: 07/15/2019 02:32:54

Diagnostics Check	OK	02:26
Air Blank	0.000	02:26
Control Test	0.081	02:27
Air Blank	0.000	02:27
Subject Sample # 1	0.169	02:28
Breath Volume	1.699	02:28
Air Blank	0.000	02:28
Air Blank	0.000	02:30
Subject Sample # 2	0.165	02:31
Breath Volume	2.753	02:31
Air Blank	0.000	02:31
Control Test	0.080	02:32
Air Blank	0.000	02:32
Diagnostics Check	OK	02:32

Data Download Date/Time: 08/03/2019 04:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA, RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: TOZZI, SHAUN
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: MARRERO

DATE/TIME OF TEST: 07/17/2019 23:23:27

Diagnostics Check	OK	23:17
Air Blank	0.000	23:17
Control Test	0.081	23:17
Air Blank	0.000	23:18
Subject Sample # 1	0.115	23:18
Breath Volume	2.886	23:18
Air Blank	0.000	23:19
Air Blank	0.000	23:21
Subject Sample # 2	0.117	23:21
Breath Volume	3.023	23:21
Air Blank	0.000	23:22
Control Test	0.081	23:22
Air Blank	0.000	23:23
Diagnostics Check	OK	23:23

Data Download Date/Time: 08/03/2019 04:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOUROUMOUSIS,GEORGE,T
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: COLLINS,SHARIKA
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 19:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:19

SUBJECT'S LAST NAME: CALVI

DATE/TIME OF TEST: 07/19/2019 20:47:34

Diagnostics Check		OK	20:41
Air Blank		0.000	20:42
Control Test		0.080	20:42
Air Blank		0.000	20:42
Subject Sample	# 1	0.224	20:43
Breath Volume		1.121	20:43
Air Blank		0.000	20:43
Air Blank		0.000	20:45
Subject Sample	# 2	0.216	20:46
Breath Volume		1.152	20:46
Air Blank		0.000	20:46
Control Test		0.078	20:46
Air Blank		0.000	20:47
Diagnostics Check		OK	20:47

Data Download Date/Time: 08/03/2019 04:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: MINICHELLO,BRITTANY
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: GUITIAN

DATE/TIME OF TEST: 07/20/2019 01:56:22

Diagnostics Check		OK	01:50
Air Blank		0.000	01:50
Control Test		0.081	01:50
Air Blank		0.000	01:51
Subject Sample	# 1	0.104	01:52
Breath Volume		1.910	01:52
Air Blank		0.000	01:52
Air Blank		0.000	01:54
Subject Sample	# 2	0.102	01:54
Breath Volume		1.675	01:54
Air Blank		0.000	01:55
Control Test		0.081	01:55
Air Blank		0.000	01:56
Diagnostics Check		OK	01:56

Data Download Date/Time: 08/03/2019 04:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOZZI,SHAUN,M
BREATH TEST OPERATOR AGENCY: Altamonte Springs PD
ARRESTING OFFICER: TOZZI,SHAUN
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 18:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: PETERSON

DATE/TIME OF TEST: 07/27/2019 19:39:03

Diagnostics Check	OK	19:32
Air Blank	0.000	19:33
Control Test	0.081	19:33
Air Blank	0.000	19:34
Subject Sample # 1	0.000	19:34
Breath Volume	2.792	19:34
Air Blank	0.000	19:35
Air Blank	0.000	19:37
Subject Sample # 2	0.000	19:37
Breath Volume	1.515	19:37
Air Blank	0.000	19:38
Control Test	0.081	19:38
Air Blank	0.000	19:38
Diagnostics Check	OK	19:39

Data Download Date/Time: 08/03/2019 04:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: HOLLAND,TROY
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: DORLUSCA

DATE/TIME OF TEST: 07/28/2019 03:58:15

Diagnostics Check	OK	03:52
Air Blank	0.000	03:52
Control Test	0.082	03:53
Air Blank	0.000	03:53
Subject Sample # 1	0.000	03:53
Breath Volume	2.789	03:53
Air Blank	0.000	03:54
Air Blank	0.000	03:56
Subject Sample # 2	0.000	03:56
Breath Volume	2.894	03:56
Air Blank	0.000	03:57
Control Test	0.078	03:57
Air Blank	0.000	03:58
Diagnostics Check	OK	03:58

Data Download Date/Time: 08/03/2019 04:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: KEE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: CAMBALL MOLINA

DATE/TIME OF TEST: 07/28/2019 04:23:57

Diagnostics Check	OK	04:15
Air Blank	0.000	04:16
Control Test	0.080	04:16
Air Blank	0.000	04:16
Subject Sample # 1	0.198	04:18
Breath Volume	1.308	04:18
Air Blank	0.000	04:18
Air Blank	0.000	04:20
Subject Sample # 2	0.206	04:22
Breath Volume	1.503	04:22
Air Blank	0.000	04:23
Control Test	0.079	04:23
Air Blank	0.000	04:23
Diagnostics Check	OK	04:23

Data Download Date/Time: 08/03/2019 04:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: QUINTAL,VIRGEL
ARRESTING OFFICER AGENCY: Longwood PD
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 03:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: NUBIN

DATE/TIME OF TEST: 07/28/2019 05:07:16

Diagnostics Check	OK	05:01
Air Blank	0.000	05:01
Control Test	0.080	05:01
Air Blank	0.000	05:02
Subject Sample # 1	0.073	05:02
Breath Volume	1.449	05:02
Air Blank	0.000	05:03
Air Blank	0.000	05:05
Subject Sample # 2	0.072	05:05
Breath Volume	1.792	05:05
Air Blank	0.000	05:06
Control Test	0.080	05:06
Air Blank	0.000	05:07
Diagnostics Check	OK	05:07

Data Download Date/Time: 08/03/2019 04:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: MASSIP,LEONARDO
ARRESTING OFFICER AGENCY: Oviedo PD
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 20:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:08

SUBJECT'S LAST NAME: ALAIDI

DATE/TIME OF TEST: 07/28/2019 21:36:33

Diagnostics Check	OK	21:28
Air Blank	0.000	21:28
Control Test	0.081	21:28
Air Blank	0.000	21:29
Subject Sample # 1	0.134	21:30
Breath Volume	1.734	21:30
Air Blank	0.000	21:31
Air Blank	0.000	21:33
Subject Sample # 2	0.126	21:34
Breath Volume	1.433	21:34
Air Blank	0.000	21:35
Control Test	0.080	21:35
Air Blank	0.000	21:36
Diagnostics Check	OK	21:36

Data Download Date/Time: 08/03/2019 04:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: HUNDEMANN,ROBERT
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: SIMONS

DATE/TIME OF TEST: 07/28/2019 22:49:33

Diagnostics Check	OK	22:42
Air Blank	0.000	22:42
Control Test	0.080	22:42
Air Blank	0.000	22:43
Subject Sample # 1	0.181	22:44
Breath Volume	1.324	22:44
Air Blank	0.000	22:45
Air Blank	0.000	22:47
Subject Sample # 2	0.184	22:48
Breath Volume	1.226	22:48
Air Blank	0.000	22:48
Control Test	0.080	22:48
Air Blank	0.000	22:49
Diagnostics Check	OK	22:49

Data Download Date/Time: 08/03/2019 04:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: KINKOPF,TIM
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: MANJI

DATE/TIME OF TEST: 07/29/2019 01:27:26

Diagnostics Check	OK	01:25	
Air Blank	0.000	01:25	
Control Test	0.080	01:25	
Air Blank	0.000	01:26	
Subject Sample	# 1	01:26	Sequence Aborted
Breath Volume	0.000	01:26	
Air Blank	0.000	01:27	

Data Download Date/Time: 08/03/2019 04:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: KINKOPF,TIM
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: MANJI

DATE/TIME OF TEST: 07/29/2019 01:38:23

Diagnostics Check	OK	01:31
Air Blank	0.000	01:31
Control Test	0.080	01:32
Air Blank	0.000	01:32
Subject Sample # 1	0.215	01:33
Breath Volume	1.691	01:33
Air Blank	0.000	01:34
Air Blank	0.000	01:35
Subject Sample # 2	0.213	01:36
Breath Volume	2.035	01:36
Air Blank	0.000	01:37
Control Test	0.080	01:37
Air Blank	0.000	01:38
Diagnostics Check	OK	01:38

Data Download Date/Time: 08/03/2019 04:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: MCKIANAN,FRANCIS
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: WADE JR

DATE/TIME OF TEST: 07/29/2019 04:49:25

Diagnostics Check		OK	04:37
Air Blank		0.000	04:37
Control Test		0.081	04:38
Air Blank		0.000	04:38
Subject Sample	# 1	0.228	04:39
Breath Volume		1.218	04:39
Air Blank		0.000	04:40
Air Blank		0.000	04:42
Subject Sample	# 2	0.252	04:44
Breath Volume		1.929	04:44
Air Blank		0.000	04:44
Air Blank		0.000	04:46
Subject Sample	# 3	0.244	04:47
Breath Volume		1.871	04:47
Air Blank		0.000	04:48
Control Test		0.078	04:48
Air Blank		0.000	04:49
Diagnostics Check		OK	04:49

Data Download Date/Time: 08/03/2019 04:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMENS,ADAM,J
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: SOTO,V
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 09:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:43

SUBJECT'S LAST NAME: CORDERO-VEGA

DATE/TIME OF TEST: 07/29/2019 12:13:25

Diagnostics Check	OK	12:04
Air Blank	0.000	12:04
Control Test	0.080	12:04
Air Blank	0.000	12:05
Subject Sample # 1	0.209	12:06
Breath Volume	2.062	12:06
Air Blank	0.000	12:07
Air Blank	0.000	12:08
Subject Sample # 2	0.190	12:11
Breath Volume	1.714	12:11
Air Blank	0.000	12:12
Control Test	0.078	12:12
Air Blank	0.000	12:13
Diagnostics Check	OK	12:13

Data Download Date/Time: 08/03/2019 04:55