# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,D,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: LOPEZ,F
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 08/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: CHAVEZ

DATE/TIME OF TEST: 08/24/2019 00:55:24

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:49 |
| Air Blank | | 0.000 | 00:49 |
| Control Test | | 0.079 | 00:49 |
| Air Blank | | 0.000 | 00:50 |
| Subject Sample | # 1 | 0.194 | 00:50 |
| Breath Volume | | 2.148 | 00:50 |
| Air Blank | | 0.000 | 00:51 |
| Air Blank | | 0.000 | 00:53 |
| Subject Sample | # 2 | 0.204 | 00:53 |
| Breath Volume | | 2.394 | 00:53 |
| Air Blank | | 0.000 | 00:54 |
| Control Test | | 0.078 | 00:54 |
| Air Blank | | 0.000 | 00:55 |
| Diagnostics Check | | OK | 00:55 |

Data Download Date/Time: 09/06/2019 09:50

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORENO,FRANCISCO,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: MORENO,FRANCISCO
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 08/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 10:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:55

SUBJECT'S LAST NAME: ESPINOZA

DATE/TIME OF TEST: 08/28/2019 11:18:47

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 11:15 | |
| Air Blank | | 0.000 | 11:16 | |
| Control Test | | 0.078 | 11:16 | |
| Air Blank | | 0.000 | 11:17 | |
| Subject Sample | # 1 | | 11:17 | Subject Test Refused |
| Breath Volume | | 0.000 | 11:17 | |
| Air Blank | | 0.000 | 11:17 | |
| Control Test | | 0.078 | 11:18 | |
| Air Blank | | 0.000 | 11:18 | |
| Diagnostics Check | | OK | 11:18 | |

Data Download Date/Time: 09/06/2019 09:50

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORENO,FRANCISCO,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: MORENO,FRANCISCO
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 08/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 11:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:22

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 08/29/2019 11:48:46

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 11:42 |
| Air Blank | | 0.000 | 11:43 |
| Control Test | | 0.078 | 11:43 |
| Air Blank | | 0.000 | 11:44 |
| Subject Sample | # 1 | 0.000 | 11:44 |
| Breath Volume | | 1.335 | 11:44 |
| Air Blank | | 0.000 | 11:45 |
| Air Blank | | 0.000 | 11:47 |
| Subject Sample | # 2 | 0.000 | 11:47 |
| Breath Volume | | 1.859 | 11:47 |
| Air Blank | | 0.000 | 11:47 |
| Control Test | | 0.077 | 11:48 |
| Air Blank | | 0.000 | 11:48 |
| Diagnostics Check | | OK | 11:48 |

Data Download Date/Time: 09/06/2019 09:50

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYRTIL,REGINALD,J
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 07/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 08/02/2019 03:33:25

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 03:30 | |
| Air Blank | | 0.000 | 03:30 | |
| Control Test | | 0.080 | 03:31 | |
| Air Blank | | 0.000 | 03:31 | |
| Subject Sample | # 1 | | 03:32 | Subject Test Refused |
| Breath Volume | | 0.000 | 03:32 | |
| Air Blank | | 0.000 | 03:32 | |
| Control Test | | 0.080 | 03:32 | |
| Air Blank | | 0.000 | 03:33 | |
| Diagnostics Check | | OK | 03:33 | |

Data Download Date/Time: 08/13/2019 10:15

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTIERREZ,KRISTINA,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: GUTIERREZ,KRISTINA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 07/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: ROJAS

DATE/TIME OF TEST: 08/04/2019 04:10:49

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 03:56 | |
| Air Blank | | 0.000 | 03:56 | |
| Control Test | | 0.080 | 03:57 | |
| Air Blank | | 0.000 | 03:57 | |
| Subject Sample | # 1 | 0.107 | 03:59 | |
| Breath Volume | | 1.777 | 03:59 | |
| Air Blank | | 0.000 | 04:00 | |
| Air Blank | | 0.000 | 04:01 | |
| Subject Sample | # 2 | 0.098 | 04:05 | Volume Not Met |
| Breath Volume | | 0.097 | 04:05 | |
| Air Blank | | 0.000 | 04:05 | |
| Air Blank | | 0.000 | 04:07 | |
| Subject Sample | # 3 | 0.098 | 04:09 | |
| Breath Volume | | 1.503 | 04:09 | |
| Air Blank | | 0.000 | 04:09 | |
| Control Test | | 0.078 | 04:10 | |
| Air Blank | | 0.000 | 04:10 | |
| Diagnostics Check | | OK | 04:10 | |

Data Download Date/Time: 08/13/2019 10:15

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAINTILIEN,BOB,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: SAINTILIEN,BOB
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 07/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: DE LA MATA

DATE/TIME OF TEST: 08/04/2019 20:52:32

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 20:48 | |
| Air Blank | | 0.000 | 20:49 | |
| Control Test | | 0.079 | 20:49 | |
| Air Blank | | 0.000 | 20:50 | |
| Subject Sample | # 1 | | 20:51 | Subject Test Refused |
| Breath Volume | | 0.000 | 20:51 | |
| Air Blank | | 0.000 | 20:51 | |
| Control Test | | 0.080 | 20:51 | |
| Air Blank | | 0.000 | 20:52 | |
| Diagnostics Check | | OK | 20:52 | |

Data Download Date/Time: 08/13/2019 10:15

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANCO,GREGORY,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 07/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:18

SUBJECT'S LAST NAME: VILCHEZ

DATE/TIME OF TEST: 08/04/2019 21:49:31

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 21:42 |
| Air Blank | | 0.000 | 21:42 |
| Control Test | | 0.080 | 21:42 |
| Air Blank | | 0.000 | 21:43 |
| Subject Sample | # 1 | 0.272 | 21:44 |
| Breath Volume | | 2.253 | 21:44 |
| Air Blank | | 0.000 | 21:45 |
| Air Blank | | 0.000 | 21:47 |
| Subject Sample | # 2 | 0.269 | 21:47 |
| Breath Volume | | 1.570 | 21:47 |
| Air Blank | | 0.000 | 21:48 |
| Control Test | | 0.078 | 21:48 |
| Air Blank | | 0.000 | 21:49 |
| Diagnostics Check | | OK | 21:49 |

Data Download Date/Time: 08/13/2019 10:15

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POVEDA,EMILIE,M
BREATH TEST OPERATOR AGENCY: SWEETWATER PD
ARRESTING OFFICER: POVEDA,EMILIE
ARRESTING OFFICER AGENCY: SWEETWATER PD
LAST AGENCY INSPECTION: 07/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: MATUS ORDENEZ

DATE/TIME OF TEST: 08/08/2019 02:09:28

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 02:04 | |
| Air Blank | | 0.000 | 02:04 | |
| Control Test | | 0.080 | 02:05 | |
| Air Blank | | 0.000 | 02:05 | |
| Subject Sample | # 1 | 0.167 | 02:08 | |
| Breath Volume | | 1.703 | 02:08 | |
| Air Blank | | | 02:09 | Purge Fail |
| Air Blank | | 0.000 | 02:09 | |

Data Download Date/Time: 08/13/2019 10:15

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POVEDA,EMILIE,M
BREATH TEST OPERATOR AGENCY: SWEETWATER PD
ARRESTING OFFICER: POVEDA,EMILIE
ARRESTING OFFICER AGENCY: SWEETWATER PD
LAST AGENCY INSPECTION: 07/11/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: MATUS ORDENEZ

DATE/TIME OF TEST: 08/08/2019 02:20:24

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:13 |
| Air Blank | | 0.000 | 02:14 |
| Control Test | | 0.078 | 02:14 |
| Air Blank | | 0.000 | 02:14 |
| Subject Sample | # 1 | 0.148 | 02:15 |
| Breath Volume | | 1.691 | 02:15 |
| Air Blank | | 0.000 | 02:16 |
| Air Blank | | 0.000 | 02:17 |
| Subject Sample | # 2 | 0.166 | 02:18 |
| Breath Volume | | 2.382 | 02:18 |
| Air Blank | | 0.000 | 02:19 |
| Control Test | | 0.079 | 02:19 |
| Air Blank | | 0.000 | 02:20 |
| Diagnostics Check | | OK | 02:20 |

Data Download Date/Time: 08/13/2019 10:15

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTANA,JORGE,F
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: QUINTANA,JORGE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 07/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 06:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:15

SUBJECT'S LAST NAME: CRUZ GARCIA

DATE/TIME OF TEST: 08/11/2019 07:43:29

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 07:35 |
| Air Blank | | 0.000 | 07:36 |
| Control Test | | 0.080 | 07:36 |
| Air Blank | | 0.000 | 07:37 |
| Subject Sample | # 1 | 0.250 | 07:38 |
| Breath Volume | | 2.503 | 07:38 |
| Air Blank | | 0.000 | 07:38 |
| Air Blank | | 0.000 | 07:40 |
| Subject Sample | # 2 | 0.245 | 07:41 |
| Breath Volume | | 1.953 | 07:41 |
| Air Blank | | 0.000 | 07:42 |
| Control Test | | 0.079 | 07:42 |
| Air Blank | | 0.000 | 07:43 |
| Diagnostics Check | | OK | 07:43 |

Data Download Date/Time: 08/13/2019 10:15

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,JAVIER,
BREATH TEST OPERATOR AGENCY: OTHER LOCAL
ARRESTING OFFICER: AGUILERA,ANDREW
ARRESTING OFFICER AGENCY: OTHER LOCAL
LAST AGENCY INSPECTION: 07/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:03

SUBJECT'S LAST NAME: CABRERA

DATE/TIME OF TEST: 08/11/2019 17:30:47

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 17:27 | |
| Air Blank | | 0.000 | 17:28 | |
| Control Test | | 0.079 | 17:28 | |
| Air Blank | | 0.000 | 17:29 | |
| Subject Sample | # 1 | | 17:29 | Subject Test Refused |
| Breath Volume | | 0.000 | 17:29 | |
| Air Blank | | 0.000 | 17:29 | |
| Control Test | | 0.080 | 17:30 | |
| Air Blank | | 0.000 | 17:30 | |
| Diagnostics Check | | OK | 17:30 | |

Data Download Date/Time: 08/13/2019 10:15

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 08/02/2019 01:36:42

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 01:33 | |
| Air Blank | | 0.000 | 01:34 | |
| Control Test | | 0.080 | 01:34 | |
| Air Blank | | 0.000 | 01:35 | |
| Subject Sample | # 1 | | 01:35 | Subject Test Refused |
| Breath Volume | | 0.000 | 01:35 | |
| Air Blank | | 0.000 | 01:35 | |
| Control Test | | 0.079 | 01:36 | |
| Air Blank | | 0.000 | 01:36 | |
| Diagnostics Check | | OK | 01:36 | |

Data Download Date/Time: 08/14/2019 11:08

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:01

SUBJECT'S LAST NAME: THARP

DATE/TIME OF TEST: 08/02/2019 05:25:53

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 05:23 | |
| Air Blank | | 0.000 | 05:23 | |
| Control Test | | 0.081 | 05:23 | |
| Air Blank | | 0.000 | 05:24 | |
| Subject Sample | # 1 | | 05:24 | Subject Test Refused |
| Breath Volume | | 0.000 | 05:24 | |
| Air Blank | | 0.000 | 05:24 | |
| Control Test | | 0.081 | 05:25 | |
| Air Blank | | 0.000 | 05:25 | |
| Diagnostics Check | | OK | 05:25 | |

Data Download Date/Time: 08/14/2019 11:08

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: SEXTON

DATE/TIME OF TEST: 08/03/2019 01:24:56

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:18 |
| Air Blank | | 0.000 | 01:18 |
| Control Test | | 0.081 | 01:19 |
| Air Blank | | 0.000 | 01:19 |
| Subject Sample | # 1 | 0.236 | 01:20 |
| Breath Volume | | 3.253 | 01:20 |
| Air Blank | | 0.000 | 01:20 |
| Air Blank | | 0.000 | 01:22 |
| Subject Sample | # 2 | 0.231 | 01:23 |
| Breath Volume | | 3.988 | 01:23 |
| Air Blank | | 0.000 | 01:24 |
| Control Test | | 0.080 | 01:24 |
| Air Blank | | 0.000 | 01:24 |
| Diagnostics Check | | OK | 01:24 |

Data Download Date/Time: 08/14/2019 11:08

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: MEDLOCK

DATE/TIME OF TEST: 08/03/2019 04:19:15

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:12 |
| Air Blank | | 0.000 | 04:13 |
| Control Test | | 0.081 | 04:13 |
| Air Blank | | 0.000 | 04:14 |
| Subject Sample | # 1 | 0.041 | 04:14 |
| Breath Volume | | 2.820 | 04:14 |
| Air Blank | | 0.000 | 04:15 |
| Air Blank | | 0.000 | 04:17 |
| Subject Sample | # 2 | 0.042 | 04:17 |
| Breath Volume | | 4.230 | 04:17 |
| Air Blank | | 0.000 | 04:18 |
| Control Test | | 0.081 | 04:18 |
| Air Blank | | 0.000 | 04:19 |
| Diagnostics Check | | OK | 04:19 |

Data Download Date/Time: 08/14/2019 11:08

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 08/04/2019 01:28:36

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 01:25 | |
| Air Blank | | 0.000 | 01:25 | |
| Control Test | | 0.081 | 01:26 | |
| Air Blank | | 0.000 | 01:26 | |
| Subject Sample | # 1 | | 01:27 | Subject Test Refused |
| Breath Volume | | 0.000 | 01:27 | |
| Air Blank | | 0.000 | 01:27 | |
| Control Test | | 0.080 | 01:28 | |
| Air Blank | | 0.000 | 01:28 | |
| Diagnostics Check | | OK | 01:28 | |

Data Download Date/Time: 08/14/2019 11:08

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEPHERD,P,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PARMENTER,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: DONALD

DATE/TIME OF TEST: 08/05/2019 18:31:54

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 18:25 |
| Air Blank | | 0.000 | 18:26 |
| Control Test | | 0.081 | 18:26 |
| Air Blank | | 0.000 | 18:27 |
| Subject Sample | # 1 | 0.000 | 18:27 |
| Breath Volume | | 2.320 | 18:27 |
| Air Blank | | 0.000 | 18:28 |
| Air Blank | | 0.000 | 18:30 |
| Subject Sample | # 2 | 0.000 | 18:30 |
| Breath Volume | | 2.316 | 18:30 |
| Air Blank | | 0.000 | 18:31 |
| Control Test | | 0.080 | 18:31 |
| Air Blank | | 0.000 | 18:31 |
| Diagnostics Check | | OK | 18:31 |

Data Download Date/Time: 08/14/2019 11:08

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DEAL,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: CASTILLO EUCEDA

DATE/TIME OF TEST: 08/08/2019 04:05:21

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:58 |
| Air Blank | | 0.000 | 03:59 |
| Control Test | | 0.080 | 03:59 |
| Air Blank | | 0.000 | 03:59 |
| Subject Sample | # 1 | 0.177 | 04:00 |
| Breath Volume | | 3.167 | 04:00 |
| Air Blank | | 0.000 | 04:01 |
| Air Blank | | 0.000 | 04:03 |
| Subject Sample | # 2 | 0.187 | 04:03 |
| Breath Volume | | 2.105 | 04:03 |
| Air Blank | | 0.000 | 04:04 |
| Control Test | | 0.081 | 04:04 |
| Air Blank | | 0.000 | 04:05 |
| Diagnostics Check | | OK | 04:05 |

Data Download Date/Time: 08/14/2019 11:08

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVERETT,I,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: OLIVER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:07

SUBJECT'S LAST NAME: COCHRAN

DATE/TIME OF TEST: 08/08/2019 18:34:01

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 18:27 |
| Air Blank | | 0.000 | 18:28 |
| Control Test | | 0.081 | 18:28 |
| Air Blank | | 0.000 | 18:29 |
| Subject Sample | # 1 | 0.000 | 18:29 |
| Breath Volume | | 1.535 | 18:29 |
| Air Blank | | 0.000 | 18:30 |
| Air Blank | | 0.000 | 18:32 |
| Subject Sample | # 2 | 0.000 | 18:32 |
| Breath Volume | | 1.246 | 18:32 |
| Air Blank | | 0.000 | 18:33 |
| Control Test | | 0.081 | 18:33 |
| Air Blank | | 0.000 | 18:33 |
| Diagnostics Check | | OK | 18:34 |

Data Download Date/Time: 08/14/2019 11:08

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FARLEY,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: MURRAH

DATE/TIME OF TEST: 08/09/2019 01:30:40

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:24 |
| Air Blank | | 0.000 | 01:24 |
| Control Test | | 0.081 | 01:25 |
| Air Blank | | 0.000 | 01:25 |
| Subject Sample | # 1 | 0.000 | 01:26 |
| Breath Volume | | 3.750 | 01:26 |
| Air Blank | | 0.000 | 01:26 |
| Air Blank | | 0.000 | 01:28 |
| Subject Sample | # 2 | 0.000 | 01:29 |
| Breath Volume | | 3.109 | 01:29 |
| Air Blank | | 0.000 | 01:29 |
| Control Test | | 0.081 | 01:30 |
| Air Blank | | 0.000 | 01:30 |
| Diagnostics Check | | OK | 01:30 |

Data Download Date/Time: 08/14/2019 11:08

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVERETT,I,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:25

SUBJECT'S LAST NAME: BATTLE

DATE/TIME OF TEST: 08/09/2019 18:54:36

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 18:45 |
| Air Blank | | 0.000 | 18:46 |
| Control Test | | 0.080 | 18:46 |
| Air Blank | | 0.000 | 18:46 |
| Subject Sample | # 1 | 0.097 | 18:49 |
| Breath Volume | | 1.285 | 18:49 |
| Air Blank | | 0.000 | 18:49 |
| Air Blank | | 0.000 | 18:51 |
| Subject Sample | # 2 | 0.087 | 18:53 |
| Breath Volume | | 1.578 | 18:53 |
| Air Blank | | 0.000 | 18:53 |
| Control Test | | 0.080 | 18:54 |
| Air Blank | | 0.000 | 18:54 |
| Diagnostics Check | | OK | 18:54 |

Data Download Date/Time: 08/14/2019 11:08

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMAS,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: KHIM

DATE/TIME OF TEST: 08/10/2019 02:00:55

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 01:58 | |
| Air Blank | | 0.000 | 01:58 | |
| Control Test | | 0.080 | 01:58 | |
| Air Blank | | 0.000 | 01:59 | |
| Subject Sample | # 1 | | 01:59 | Subject Test Refused |
| Breath Volume | | 0.000 | 01:59 | |
| Air Blank | | 0.000 | 02:00 | |
| Control Test | | 0.080 | 02:00 | |
| Air Blank | | 0.000 | 02:00 | |
| Diagnostics Check | | OK | 02:00 | |

Data Download Date/Time: 08/14/2019 11:08

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JENSEN,D
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: VARGAS

DATE/TIME OF TEST: 08/10/2019 03:24:21

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 03:21 | |
| Air Blank | | 0.000 | 03:21 | |
| Control Test | | 0.080 | 03:22 | |
| Air Blank | | 0.000 | 03:22 | |
| Subject Sample | # 1 | | 03:23 | Subject Test Refused |
| Breath Volume | | 0.000 | 03:23 | |
| Air Blank | | 0.000 | 03:23 | |
| Control Test | | 0.080 | 03:23 | |
| Air Blank | | 0.000 | 03:24 | |
| Diagnostics Check | | OK | 03:24 | |

Data Download Date/Time: 08/14/2019 11:08

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 08/10/2019 23:27:59

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 23:25 | |
| Air Blank | | 0.000 | 23:25 | |
| Control Test | | 0.081 | 23:25 | |
| Air Blank | | 0.000 | 23:26 | |
| Subject Sample | # 1 | | 23:26 | Subject Test Refused |
| Breath Volume | | 0.000 | 23:26 | |
| Air Blank | | 0.000 | 23:27 | |
| Control Test | | 0.081 | 23:27 | |
| Air Blank | | 0.000 | 23:27 | |
| Diagnostics Check | | OK | 23:27 | |

Data Download Date/Time: 08/14/2019 11:08

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 08/11/2019 02:14:49

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:07 |
| Air Blank | | 0.000 | 02:07 |
| Control Test | | 0.081 | 02:07 |
| Air Blank | | 0.000 | 02:08 |
| Subject Sample | # 1 | 0.263 | 02:08 |
| Breath Volume | | 1.746 | 02:08 |
| Air Blank | | 0.000 | 02:09 |
| Air Blank | | 0.000 | 02:11 |
| Subject Sample | # 2 | 0.245 | 02:13 |
| Breath Volume | | 1.937 | 02:13 |
| Air Blank | | 0.000 | 02:13 |
| Control Test | | 0.081 | 02:14 |
| Air Blank | | 0.000 | 02:14 |
| Diagnostics Check | | OK | 02:14 |

Data Download Date/Time: 08/14/2019 11:08

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DIAZ,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: VELASQUEZ JIMENEZ

DATE/TIME OF TEST: 08/11/2019 04:45:56

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:39 |
| Air Blank | | 0.000 | 04:40 |
| Control Test | | 0.081 | 04:40 |
| Air Blank | | 0.000 | 04:40 |
| Subject Sample | # 1 | 0.214 | 04:41 |
| Breath Volume | | 2.703 | 04:41 |
| Air Blank | | 0.000 | 04:42 |
| Air Blank | | 0.000 | 04:43 |
| Subject Sample | # 2 | 0.207 | 04:44 |
| Breath Volume | | 1.882 | 04:44 |
| Air Blank | | 0.000 | 04:45 |
| Control Test | | 0.080 | 04:45 |
| Air Blank | | 0.000 | 04:45 |
| Diagnostics Check | | OK | 04:45 |

Data Download Date/Time: 08/14/2019 11:08

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVERETT,I,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MAYS,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 06:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:34

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 08/11/2019 08:01:24

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 07:54 |
| Air Blank | | 0.000 | 07:54 |
| Control Test | | 0.081 | 07:55 |
| Air Blank | | 0.000 | 07:55 |
| Subject Sample | # 1 | 0.217 | 07:56 |
| Breath Volume | | 1.113 | 07:56 |
| Air Blank | | 0.000 | 07:57 |
| Air Blank | | 0.000 | 07:59 |
| Subject Sample | # 2 | 0.202 | 07:59 |
| Breath Volume | | 2.085 | 07:59 |
| Air Blank | | 0.000 | 08:00 |
| Control Test | | 0.081 | 08:00 |
| Air Blank | | 0.000 | 08:01 |
| Diagnostics Check | | OK | 08:01 |

Data Download Date/Time: 08/14/2019 11:08

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MCFARLAND,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: SEBASTIAN

DATE/TIME OF TEST: 08/12/2019 03:01:02

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 02:58 | |
| Air Blank | | 0.000 | 02:58 | |
| Control Test | | 0.081 | 02:58 | |
| Air Blank | | 0.000 | 02:59 | |
| Subject Sample | # 1 | | 02:59 | Subject Test Refused |
| Breath Volume | | 0.000 | 02:59 | |
| Air Blank | | 0.000 | 03:00 | |
| Control Test | | 0.080 | 03:00 | |
| Air Blank | | 0.000 | 03:00 | |
| Diagnostics Check | | OK | 03:01 | |

Data Download Date/Time: 08/14/2019 11:08

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: NEWSON

DATE/TIME OF TEST: 08/01/2019 03:55:50

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:48 |
| Air Blank | | 0.000 | 03:48 |
| Control Test | | 0.080 | 03:49 |
| Air Blank | | 0.000 | 03:49 |
| Subject Sample | # 1 | 0.148 | 03:50 |
| Breath Volume | | 1.300 | 03:50 |
| Air Blank | | 0.000 | 03:50 |
| Air Blank | | 0.000 | 03:52 |
| Subject Sample | # 2 | 0.140 | 03:54 |
| Breath Volume | | 2.257 | 03:54 |
| Air Blank | | 0.000 | 03:54 |
| Control Test | | 0.079 | 03:55 |
| Air Blank | | 0.000 | 03:55 |
| Diagnostics Check | | OK | 03:55 |

Data Download Date/Time: 08/14/2019 11:21

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: LYMAN

DATE/TIME OF TEST: 08/02/2019 23:23:08

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:16 |
| Air Blank | | 0.000 | 23:17 |
| Control Test | | 0.079 | 23:17 |
| Air Blank | | 0.000 | 23:17 |
| Subject Sample | # 1 | 0.073 | 23:18 |
| Breath Volume | | 1.531 | 23:18 |
| Air Blank | | 0.000 | 23:19 |
| Air Blank | | 0.000 | 23:21 |
| Subject Sample | # 2 | 0.067 | 23:21 |
| Breath Volume | | 1.828 | 23:21 |
| Air Blank | | 0.000 | 23:22 |
| Control Test | | 0.079 | 23:22 |
| Air Blank | | 0.000 | 23:23 |
| Diagnostics Check | | OK | 23:23 |

Data Download Date/Time: 08/14/2019 11:21

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 08/03/2019 01:10:24

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:03 |
| Air Blank | | 0.000 | 01:04 |
| Control Test | | 0.079 | 01:04 |
| Air Blank | | 0.000 | 01:05 |
| Subject Sample | # 1 | 0.192 | 01:05 |
| Breath Volume | | 2.390 | 01:05 |
| Air Blank | | 0.000 | 01:06 |
| Air Blank | | 0.000 | 01:08 |
| Subject Sample | # 2 | 0.187 | 01:08 |
| Breath Volume | | 2.093 | 01:08 |
| Air Blank | | 0.000 | 01:09 |
| Control Test | | 0.078 | 01:09 |
| Air Blank | | 0.000 | 01:10 |
| Diagnostics Check | | OK | 01:10 |

Data Download Date/Time: 08/14/2019 11:21

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TIMMONS,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: KOZLOSKI

DATE/TIME OF TEST: 08/03/2019 03:44:10

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:37 |
| Air Blank | | 0.000 | 03:38 |
| Control Test | | 0.079 | 03:38 |
| Air Blank | | 0.000 | 03:38 |
| Subject Sample | # 1 | 0.121 | 03:39 |
| Breath Volume | | 2.054 | 03:39 |
| Air Blank | | 0.000 | 03:40 |
| Air Blank | | 0.000 | 03:42 |
| Subject Sample | # 2 | 0.122 | 03:42 |
| Breath Volume | | 2.113 | 03:42 |
| Air Blank | | 0.000 | 03:43 |
| Control Test | | 0.079 | 03:43 |
| Air Blank | | 0.000 | 03:44 |
| Diagnostics Check | | OK | 03:44 |

Data Download Date/Time: 08/14/2019 11:21

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:14

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 08/03/2019 06:40:51

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 06:34 |
| Air Blank | | 0.000 | 06:35 |
| Control Test | | 0.079 | 06:35 |
| Air Blank | | 0.000 | 06:36 |
| Subject Sample | # 1 | 0.134 | 06:36 |
| Breath Volume | | 2.433 | 06:36 |
| Air Blank | | 0.000 | 06:37 |
| Air Blank | | 0.000 | 06:38 |
| Subject Sample | # 2 | 0.130 | 06:39 |
| Breath Volume | | 3.140 | 06:39 |
| Air Blank | | 0.000 | 06:39 |
| Control Test | | 0.079 | 06:40 |
| Air Blank | | 0.000 | 06:40 |
| Diagnostics Check | | OK | 06:40 |

Data Download Date/Time: 08/14/2019 11:21

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 08/04/2019 00:41:18

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:35 |
| Air Blank | | 0.000 | 00:35 |
| Control Test | | 0.079 | 00:36 |
| Air Blank | | 0.000 | 00:36 |
| Subject Sample | # 1 | 0.155 | 00:36 |
| Breath Volume | | 2.664 | 00:36 |
| Air Blank | | 0.000 | 00:37 |
| Air Blank | | 0.000 | 00:39 |
| Subject Sample | # 2 | 0.159 | 00:39 |
| Breath Volume | | 2.593 | 00:39 |
| Air Blank | | 0.000 | 00:40 |
| Control Test | | 0.079 | 00:40 |
| Air Blank | | 0.000 | 00:41 |
| Diagnostics Check | | OK | 00:41 |

Data Download Date/Time: 08/14/2019 11:21

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: SLOAN

DATE/TIME OF TEST: 08/04/2019 04:56:54

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:50 |
| Air Blank | | 0.000 | 04:50 |
| Control Test | | 0.079 | 04:51 |
| Air Blank | | 0.000 | 04:51 |
| Subject Sample | # 1 | 0.204 | 04:52 |
| Breath Volume | | 2.582 | 04:52 |
| Air Blank | | 0.000 | 04:52 |
| Air Blank | | 0.000 | 04:54 |
| Subject Sample | # 2 | 0.197 | 04:55 |
| Breath Volume | | 2.753 | 04:55 |
| Air Blank | | 0.000 | 04:55 |
| Control Test | | 0.079 | 04:56 |
| Air Blank | | 0.000 | 04:56 |
| Diagnostics Check | | OK | 04:56 |

Data Download Date/Time: 08/14/2019 11:21

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: CALLE MURILLO

DATE/TIME OF TEST: 08/05/2019 00:20:15

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:13 |
| Air Blank | | 0.000 | 00:14 |
| Control Test | | 0.080 | 00:14 |
| Air Blank | | 0.000 | 00:14 |
| Subject Sample | # 1 | 0.181 | 00:15 |
| Breath Volume | | 1.464 | 00:15 |
| Air Blank | | 0.000 | 00:16 |
| Air Blank | | 0.000 | 00:17 |
| Subject Sample | # 2 | 0.188 | 00:18 |
| Breath Volume | | 1.601 | 00:18 |
| Air Blank | | 0.000 | 00:19 |
| Control Test | | 0.079 | 00:19 |
| Air Blank | | 0.000 | 00:20 |
| Diagnostics Check | | OK | 00:20 |

Data Download Date/Time: 08/14/2019 11:21

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: BUCKLES

DATE/TIME OF TEST: 08/08/2019 02:02:22

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 01:59 | |
| Air Blank | | 0.000 | 02:00 | |
| Control Test | | 0.079 | 02:00 | |
| Air Blank | | 0.000 | 02:00 | |
| Subject Sample | # 1 | | 02:01 | Subject Test Refused |
| Breath Volume | | 0.000 | 02:01 | |
| Air Blank | | 0.000 | 02:01 | |
| Control Test | | 0.079 | 02:01 | |
| Air Blank | | 0.000 | 02:02 | |
| Diagnostics Check | | OK | 02:02 | |

Data Download Date/Time: 08/14/2019 11:21

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KAMPFE,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 08/08/2019 04:28:11

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 04:25 | |
| Air Blank | | 0.000 | 04:25 | |
| Control Test | | 0.080 | 04:26 | |
| Air Blank | | 0.000 | 04:26 | |
| Subject Sample | # 1 | | 04:26 | Subject Test Refused |
| Breath Volume | | 0.000 | 04:26 | |
| Air Blank | | 0.000 | 04:27 | |
| Control Test | | 0.080 | 04:27 | |
| Air Blank | | 0.000 | 04:28 | |
| Diagnostics Check | | OK | 04:28 | |

Data Download Date/Time: 08/14/2019 11:21

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:32

SUBJECT'S LAST NAME: CULP

DATE/TIME OF TEST: 08/08/2019 20:55:01

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 20:52 | |
| Air Blank | | 0.000 | 20:52 | |
| Control Test | | 0.080 | 20:52 | |
| Air Blank | | 0.000 | 20:53 | |
| Subject Sample | # 1 | | 20:53 | Subject Test Refused |
| Breath Volume | | 0.000 | 20:53 | |
| Air Blank | | 0.000 | 20:54 | |
| Control Test | | 0.079 | 20:54 | |
| Air Blank | | 0.000 | 20:54 | |
| Diagnostics Check | | OK | 20:55 | |

Data Download Date/Time: 08/14/2019 11:21

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TIMMONS,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: BRAMUCHI

DATE/TIME OF TEST: 08/09/2019 03:05:24

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 03:02 | |
| Air Blank | | 0.000 | 03:02 | |
| Control Test | | 0.080 | 03:03 | |
| Air Blank | | 0.000 | 03:03 | |
| Subject Sample | # 1 | | 03:04 | Subject Test Refused |
| Breath Volume | | 0.000 | 03:04 | |
| Air Blank | | 0.000 | 03:04 | |
| Control Test | | 0.079 | 03:04 | |
| Air Blank | | 0.000 | 03:05 | |
| Diagnostics Check | | OK | 03:05 | |

Data Download Date/Time: 08/14/2019 11:21

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FILBERT,E
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:33

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 08/09/2019 04:59:45

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:53 |
| Air Blank | | 0.000 | 04:53 |
| Control Test | | 0.080 | 04:53 |
| Air Blank | | 0.000 | 04:54 |
| Subject Sample | # 1 | 0.161 | 04:55 |
| Breath Volume | | 1.230 | 04:55 |
| Air Blank | | 0.000 | 04:55 |
| Air Blank | | 0.000 | 04:57 |
| Subject Sample | # 2 | 0.164 | 04:58 |
| Breath Volume | | 1.257 | 04:58 |
| Air Blank | | 0.000 | 04:58 |
| Control Test | | 0.079 | 04:59 |
| Air Blank | | 0.000 | 04:59 |
| Diagnostics Check | | OK | 04:59 |

Data Download Date/Time: 08/14/2019 11:21

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JORDAN,A
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: BAITY

DATE/TIME OF TEST: 08/10/2019 01:22:09

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:14 |
| Air Blank | | 0.000 | 01:15 |
| Control Test | | 0.080 | 01:15 |
| Air Blank | | 0.000 | 01:16 |
| Subject Sample | # 1 | 0.138 | 01:17 |
| Breath Volume | | 2.273 | 01:17 |
| Air Blank | | 0.000 | 01:17 |
| Air Blank | | 0.000 | 01:19 |
| Subject Sample | # 2 | 0.134 | 01:20 |
| Breath Volume | | 2.015 | 01:20 |
| Air Blank | | 0.000 | 01:21 |
| Control Test | | 0.078 | 01:21 |
| Air Blank | | 0.000 | 01:22 |
| Diagnostics Check | | OK | 01:22 |

Data Download Date/Time: 08/14/2019 11:21

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: KENNEDY

DATE/TIME OF TEST: 08/10/2019 03:24:20

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:12 |
| Air Blank | | 0.000 | 03:13 |
| Control Test | | 0.080 | 03:13 |
| Air Blank | | 0.000 | 03:13 |
| Subject Sample | # 1 | 0.193 | 03:15 |
| Breath Volume | | 2.226 | 03:15 |
| Air Blank | | 0.000 | 03:16 |
| Air Blank | | 0.000 | 03:18 |
| Subject Sample | # 2 | 0.215 | 03:19 |
| Breath Volume | | 4.578 | 03:19 |
| Air Blank | | 0.000 | 03:20 |
| Air Blank | | 0.000 | 03:21 |
| Subject Sample | # 3 | 0.177 | 03:22 |
| Breath Volume | | 2.070 | 03:22 |
| Air Blank | | 0.000 | 03:23 |
| Control Test | | 0.078 | 03:23 |
| Air Blank | | 0.000 | 03:24 |
| Diagnostics Check | | OK | 03:24 |

Data Download Date/Time: 08/14/2019 11:21

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: GARNES

DATE/TIME OF TEST: 08/10/2019 21:29:12

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 21:20 |
| Air Blank | | 0.000 | 21:21 |
| Control Test | | 0.080 | 21:21 |
| Air Blank | | 0.000 | 21:22 |
| Subject Sample | # 1 | 0.170 | 21:24 |
| Breath Volume | | 2.511 | 21:24 |
| Air Blank | | 0.000 | 21:24 |
| Air Blank | | 0.000 | 21:26 |
| Subject Sample | # 2 | 0.178 | 21:27 |
| Breath Volume | | 1.585 | 21:27 |
| Air Blank | | 0.000 | 21:28 |
| Control Test | | 0.079 | 21:28 |
| Air Blank | | 0.000 | 21:29 |
| Diagnostics Check | | OK | 21:29 |

Data Download Date/Time: 08/14/2019 11:21

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: SAFI

DATE/TIME OF TEST: 08/11/2019 01:40:39

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:33 |
| Air Blank | | 0.000 | 01:34 |
| Control Test | | 0.079 | 01:34 |
| Air Blank | | 0.000 | 01:34 |
| Subject Sample | # 1 | 0.020 | 01:36 |
| Breath Volume | | 3.496 | 01:36 |
| Air Blank | | 0.000 | 01:36 |
| Air Blank | | 0.000 | 01:38 |
| Subject Sample | # 2 | 0.018 | 01:39 |
| Breath Volume | | 2.898 | 01:39 |
| Air Blank | | 0.000 | 01:39 |
| Control Test | | 0.079 | 01:40 |
| Air Blank | | 0.000 | 01:40 |
| Diagnostics Check | | OK | 01:40 |

Data Download Date/Time: 08/14/2019 11:21

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHERENA,F
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: CROWE

DATE/TIME OF TEST: 08/11/2019 03:19:29

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:12 |
| Air Blank | | 0.000 | 03:12 |
| Control Test | | 0.079 | 03:13 |
| Air Blank | | 0.000 | 03:13 |
| Subject Sample | # 1 | 0.191 | 03:14 |
| Breath Volume | | 3.371 | 03:14 |
| Air Blank | | 0.000 | 03:15 |
| Air Blank | | 0.000 | 03:16 |
| Subject Sample | # 2 | 0.189 | 03:17 |
| Breath Volume | | 3.773 | 03:17 |
| Air Blank | | 0.000 | 03:18 |
| Control Test | | 0.079 | 03:18 |
| Air Blank | | 0.000 | 03:19 |
| Diagnostics Check | | OK | 03:19 |

Data Download Date/Time: 08/14/2019 11:21

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: CHAPLIN

DATE/TIME OF TEST: 08/12/2019 02:29:06

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 02:26 | |
| Air Blank | | 0.000 | 02:26 | |
| Control Test | | 0.080 | 02:26 | |
| Air Blank | | 0.000 | 02:27 | |
| Subject Sample | # 1 | | 02:27 | Subject Test Refused |
| Breath Volume | | 0.000 | 02:27 | |
| Air Blank | | 0.000 | 02:28 | |
| Control Test | | 0.080 | 02:28 | |
| Air Blank | | 0.000 | 02:29 | |
| Diagnostics Check | | OK | 02:29 | |

Data Download Date/Time: 08/14/2019 11:21

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: KRAUSE

DATE/TIME OF TEST: 08/13/2019 00:06:11

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:59 |
| Air Blank | | 0.000 | 23:59 |
| Control Test | | 0.080 | 23:59 |
| Air Blank | | 0.000 | 00:00 |
| Subject Sample | # 1 | 0.139 | 00:01 |
| Breath Volume | | 3.007 | 00:01 |
| Air Blank | | 0.000 | 00:02 |
| Air Blank | | 0.000 | 00:03 |
| Subject Sample | # 2 | 0.145 | 00:04 |
| Breath Volume | | 2.964 | 00:04 |
| Air Blank | | 0.000 | 00:05 |
| Control Test | | 0.079 | 00:05 |
| Air Blank | | 0.000 | 00:06 |
| Diagnostics Check | | OK | 00:06 |

Data Download Date/Time: 08/14/2019 11:21

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 08/13/2019 03:45:01

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 03:41 | |
| Air Blank | | 0.000 | 03:42 | |
| Control Test | | 0.080 | 03:42 | |
| Air Blank | | 0.000 | 03:43 | |
| Subject Sample | # 1 | | 03:43 | Subject Test Refused |
| Breath Volume | | 0.000 | 03:43 | |
| Air Blank | | 0.000 | 03:44 | |
| Control Test | | 0.080 | 03:44 | |
| Air Blank | | 0.000 | 03:44 | |
| Diagnostics Check | | OK | 03:45 | |

Data Download Date/Time: 08/14/2019 11:21

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: SANCHEZ ERAZO

DATE/TIME OF TEST: 08/18/2019 00:52:29

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 00:49 | |
| Air Blank | | 0.000 | 00:50 | |
| Control Test | | 0.079 | 00:50 | |
| Air Blank | | 0.000 | 00:50 | |
| Subject Sample | # 1 | | 00:51 | Subject Test Refused |
| Breath Volume | | 0.000 | 00:51 | |
| Air Blank | | 0.000 | 00:51 | |
| Control Test | | 0.080 | 00:51 | |
| Air Blank | | 0.000 | 00:52 | |
| Diagnostics Check | | OK | 00:52 | |

Data Download Date/Time: 08/30/2019 10:41

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 08/18/2019 02:08:58

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 02:06 | |
| Air Blank | | 0.000 | 02:06 | |
| Control Test | | 0.079 | 02:06 | |
| Air Blank | | 0.000 | 02:07 | |
| Subject Sample | # 1 | | 02:07 | Subject Test Refused |
| Breath Volume | | 0.000 | 02:07 | |
| Air Blank | | 0.000 | 02:08 | |
| Control Test | | 0.079 | 02:08 | |
| Air Blank | | 0.000 | 02:08 | |
| Diagnostics Check | | OK | 02:08 | |

Data Download Date/Time: 08/30/2019 10:41

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: WEST

DATE/TIME OF TEST: 08/18/2019 03:11:55

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 03:09 | |
| Air Blank | | 0.000 | 03:09 | |
| Control Test | | 0.079 | 03:09 | |
| Air Blank | | 0.000 | 03:10 | |
| Subject Sample | # 1 | | 03:10 | Subject Test Refused |
| Breath Volume | | 0.000 | 03:10 | |
| Air Blank | | 0.000 | 03:11 | |
| Control Test | | 0.080 | 03:11 | |
| Air Blank | | 0.000 | 03:11 | |
| Diagnostics Check | | OK | 03:11 | |

Data Download Date/Time: 08/30/2019 10:41

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: CURRY

DATE/TIME OF TEST: 08/18/2019 04:03:37

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 04:00 | |
| Air Blank | | 0.000 | 04:01 | |
| Control Test | | 0.080 | 04:01 | |
| Air Blank | | 0.000 | 04:01 | |
| Subject Sample | # 1 | | 04:02 | Subject Test Refused |
| Breath Volume | | 0.000 | 04:02 | |
| Air Blank | | 0.000 | 04:02 | |
| Control Test | | 0.079 | 04:03 | |
| Air Blank | | 0.000 | 04:03 | |
| Diagnostics Check | | OK | 04:03 | |

Data Download Date/Time: 08/30/2019 10:41

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOROZOV,S
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: DENNARD

DATE/TIME OF TEST: 08/18/2019 04:31:00

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 04:28 | |
| Air Blank | | 0.000 | 04:28 | |
| Control Test | | 0.079 | 04:28 | |
| Air Blank | | 0.000 | 04:29 | |
| Subject Sample | # 1 | | 04:29 | Subject Test Refused |
| Breath Volume | | 0.000 | 04:29 | |
| Air Blank | | 0.000 | 04:30 | |
| Control Test | | 0.079 | 04:30 | |
| Air Blank | | 0.000 | 04:30 | |
| Diagnostics Check | | OK | 04:31 | |

Data Download Date/Time: 08/30/2019 10:41

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PEREZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:23

SUBJECT'S LAST NAME: SINORACKI

DATE/TIME OF TEST: 08/18/2019 05:46:52

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 05:44 | |
| Air Blank | | 0.000 | 05:44 | |
| Control Test | | 0.080 | 05:44 | |
| Air Blank | | 0.000 | 05:45 | |
| Subject Sample | # 1 | | 05:45 | Subject Test Refused |
| Breath Volume | | 0.000 | 05:45 | |
| Air Blank | | 0.000 | 05:45 | |
| Control Test | | 0.080 | 05:46 | |
| Air Blank | | 0.000 | 05:46 | |
| Diagnostics Check | | OK | 05:46 | |

Data Download Date/Time: 08/30/2019 10:41

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: SCANLON

DATE/TIME OF TEST: 08/18/2019 23:26:26

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:20 |
| Air Blank | | 0.000 | 23:20 |
| Control Test | | 0.080 | 23:20 |
| Air Blank | | 0.000 | 23:21 |
| Subject Sample | # 1 | 0.161 | 23:21 |
| Breath Volume | | 2.355 | 23:21 |
| Air Blank | | 0.000 | 23:22 |
| Air Blank | | 0.000 | 23:24 |
| Subject Sample | # 2 | 0.155 | 23:24 |
| Breath Volume | | 2.382 | 23:24 |
| Air Blank | | 0.000 | 23:25 |
| Control Test | | 0.078 | 23:25 |
| Air Blank | | 0.000 | 23:26 |
| Diagnostics Check | | OK | 23:26 |

Data Download Date/Time: 08/30/2019 10:41

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DEAN,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: POLLARD III

DATE/TIME OF TEST: 08/19/2019 01:50:49

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 01:48 | |
| Air Blank | | 0.000 | 01:48 | |
| Control Test | | 0.080 | 01:48 | |
| Air Blank | | 0.000 | 01:49 | |
| Subject Sample | # 1 | | 01:49 | Subject Test Refused |
| Breath Volume | | 0.000 | 01:49 | |
| Air Blank | | 0.000 | 01:49 | |
| Control Test | | 0.079 | 01:50 | |
| Air Blank | | 0.000 | 01:50 | |
| Diagnostics Check | | OK | 01:50 | |

Data Download Date/Time: 08/30/2019 10:41

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: NA,NA
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: ADMIN
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 08/19/2019 23:12:14

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:05 |
| Air Blank | | 0.000 | 23:06 |
| Control Test | | 0.080 | 23:06 |
| Air Blank | | 0.000 | 23:07 |
| Subject Sample | # 1 | 0.078 | 23:07 |
| Breath Volume | | 1.804 | 23:07 |
| Air Blank | | 0.000 | 23:08 |
| Air Blank | | 0.000 | 23:10 |
| Subject Sample | # 2 | 0.080 | 23:10 |
| Breath Volume | | 2.941 | 23:10 |
| Air Blank | | 0.000 | 23:11 |
| Control Test | | 0.080 | 23:11 |
| Air Blank | | 0.000 | 23:12 |
| Diagnostics Check | | OK | 23:12 |

Data Download Date/Time: 08/30/2019 10:41

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MORGAN,TROOPER
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: MCMASTERS

DATE/TIME OF TEST: 08/20/2019 04:15:43

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 04:12 | |
| Air Blank | | 0.000 | 04:13 | |
| Control Test | | 0.080 | 04:13 | |
| Air Blank | | 0.000 | 04:14 | |
| Subject Sample | # 1 | | 04:14 | Subject Test Refused |
| Breath Volume | | 0.000 | 04:14 | |
| Air Blank | | 0.000 | 04:14 | |
| Control Test | | 0.080 | 04:15 | |
| Air Blank | | 0.000 | 04:15 | |
| Diagnostics Check | | OK | 04:15 | |

Data Download Date/Time: 08/30/2019 10:41

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:36

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 08/20/2019 20:03:39

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 19:56 |
| Air Blank | | 0.000 | 19:57 |
| Control Test | | 0.080 | 19:57 |
| Air Blank | | 0.000 | 19:57 |
| Subject Sample | # 1 | 0.162 | 19:59 |
| Breath Volume | | 3.667 | 19:59 |
| Air Blank | | 0.000 | 19:59 |
| Air Blank | | 0.000 | 20:01 |
| Subject Sample | # 2 | 0.165 | 20:02 |
| Breath Volume | | 3.195 | 20:02 |
| Air Blank | | 0.000 | 20:02 |
| Control Test | | 0.079 | 20:03 |
| Air Blank | | 0.000 | 20:03 |
| Diagnostics Check | | OK | 20:03 |

Data Download Date/Time: 08/30/2019 10:41

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: HERRERA

DATE/TIME OF TEST: 08/21/2019 03:55:40

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 03:52 | |
| Air Blank | | 0.000 | 03:53 | |
| Control Test | | 0.080 | 03:53 | |
| Air Blank | | 0.000 | 03:54 | |
| Subject Sample | # 1 | | 03:54 | Subject Test Refused |
| Breath Volume | | 0.000 | 03:54 | |
| Air Blank | | 0.000 | 03:54 | |
| Control Test | | 0.079 | 03:55 | |
| Air Blank | | 0.000 | 03:55 | |
| Diagnostics Check | | OK | 03:55 | |

Data Download Date/Time: 08/30/2019 10:41

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: PERKINS

DATE/TIME OF TEST: 08/22/2019 00:21:07

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:14 |
| Air Blank | | 0.000 | 00:15 |
| Control Test | | 0.079 | 00:15 |
| Air Blank | | 0.000 | 00:15 |
| Subject Sample | # 1 | 0.168 | 00:16 |
| Breath Volume | | 1.453 | 00:16 |
| Air Blank | | 0.000 | 00:17 |
| Air Blank | | 0.000 | 00:18 |
| Subject Sample | # 2 | 0.167 | 00:19 |
| Breath Volume | | 2.570 | 00:19 |
| Air Blank | | 0.000 | 00:20 |
| Control Test | | 0.079 | 00:20 |
| Air Blank | | 0.000 | 00:21 |
| Diagnostics Check | | OK | 00:21 |

Data Download Date/Time: 08/30/2019 10:41

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DEEL,T
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: RUKAB

DATE/TIME OF TEST: 08/24/2019 02:08:32

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:02 |
| Air Blank | | 0.000 | 02:03 |
| Control Test | | 0.079 | 02:03 |
| Air Blank | | 0.000 | 02:03 |
| Subject Sample | # 1 | 0.082 | 02:04 |
| Breath Volume | | 2.492 | 02:04 |
| Air Blank | | 0.000 | 02:04 |
| Air Blank | | 0.000 | 02:06 |
| Subject Sample | # 2 | 0.078 | 02:07 |
| Breath Volume | | 1.582 | 02:07 |
| Air Blank | | 0.000 | 02:07 |
| Control Test | | 0.079 | 02:07 |
| Air Blank | | 0.000 | 02:08 |
| Diagnostics Check | | OK | 02:08 |

Data Download Date/Time: 08/30/2019 10:41

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DEEL,T
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:38

SUBJECT'S LAST NAME: MCDONOUGH

DATE/TIME OF TEST: 08/24/2019 06:04:57

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 05:58 |
| Air Blank | | 0.000 | 05:59 |
| Control Test | | 0.079 | 05:59 |
| Air Blank | | 0.000 | 06:00 |
| Subject Sample | # 1 | 0.168 | 06:00 |
| Breath Volume | | 1.609 | 06:00 |
| Air Blank | | 0.000 | 06:01 |
| Air Blank | | 0.000 | 06:02 |
| Subject Sample | # 2 | 0.168 | 06:03 |
| Breath Volume | | 2.152 | 06:03 |
| Air Blank | | 0.000 | 06:04 |
| Control Test | | 0.078 | 06:04 |
| Air Blank | | 0.000 | 06:04 |
| Diagnostics Check | | OK | 06:04 |

Data Download Date/Time: 08/30/2019 10:41

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: LANDEROS

DATE/TIME OF TEST: 08/24/2019 23:06:35

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 22:59 |
| Air Blank | | 0.000 | 22:59 |
| Control Test | | 0.079 | 23:00 |
| Air Blank | | 0.000 | 23:00 |
| Subject Sample | # 1 | 0.221 | 23:01 |
| Breath Volume | | 1.453 | 23:01 |
| Air Blank | | 0.000 | 23:01 |
| Air Blank | | 0.000 | 23:03 |
| Subject Sample | # 2 | 0.229 | 23:04 |
| Breath Volume | | 1.105 | 23:04 |
| Air Blank | | 0.000 | 23:05 |
| Control Test | | 0.079 | 23:05 |
| Air Blank | | 0.000 | 23:06 |
| Diagnostics Check | | OK | 23:06 |

Data Download Date/Time: 08/30/2019 10:41

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: MARCUS

DATE/TIME OF TEST: 08/25/2019 01:16:30

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:07 |
| Air Blank | | 0.000 | 01:08 |
| Control Test | | 0.079 | 01:08 |
| Air Blank | | 0.000 | 01:09 |
| Subject Sample | # 1 | 0.155 | 01:09 |
| Breath Volume | | 1.871 | 01:09 |
| Air Blank | | 0.000 | 01:10 |
| Air Blank | | 0.000 | 01:12 |
| Subject Sample | # 2 | 0.144 | 01:14 |
| Breath Volume | | 1.296 | 01:14 |
| Air Blank | | 0.000 | 01:15 |
| Control Test | | 0.079 | 01:15 |
| Air Blank | | 0.000 | 01:16 |
| Diagnostics Check | | OK | 01:16 |

Data Download Date/Time: 08/30/2019 10:41

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LORD,B
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: ANTHONY

DATE/TIME OF TEST: 08/25/2019 04:58:57

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:52 |
| Air Blank | | 0.000 | 04:53 |
| Control Test | | 0.079 | 04:53 |
| Air Blank | | 0.000 | 04:54 |
| Subject Sample | # 1 | 0.141 | 04:54 |
| Breath Volume | | 2.593 | 04:54 |
| Air Blank | | 0.000 | 04:55 |
| Air Blank | | 0.000 | 04:57 |
| Subject Sample | # 2 | 0.137 | 04:57 |
| Breath Volume | | 2.984 | 04:57 |
| Air Blank | | 0.000 | 04:58 |
| Control Test | | 0.079 | 04:58 |
| Air Blank | | 0.000 | 04:58 |
| Diagnostics Check | | OK | 04:58 |

Data Download Date/Time: 08/30/2019 10:41

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 08/26/2019 02:44:10

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:38 |
| Air Blank | | 0.000 | 02:38 |
| Control Test | | 0.079 | 02:38 |
| Air Blank | | 0.000 | 02:39 |
| Subject Sample | # 1 | 0.240 | 02:39 |
| Breath Volume | | 2.808 | 02:39 |
| Air Blank | | 0.000 | 02:40 |
| Air Blank | | 0.000 | 02:42 |
| Subject Sample | # 2 | 0.234 | 02:42 |
| Breath Volume | | 3.164 | 02:42 |
| Air Blank | | 0.000 | 02:43 |
| Control Test | | 0.078 | 02:43 |
| Air Blank | | 0.000 | 02:44 |
| Diagnostics Check | | OK | 02:44 |

Data Download Date/Time: 08/30/2019 10:41

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 08/27/2019 01:48:43

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 01:40 | |
| Air Blank | | 0.000 | 01:40 | |
| Control Test | | 0.080 | 01:41 | |
| Air Blank | | 0.000 | 01:41 | |
| Subject Sample | # 1 | 0.175 | 01:44 | Slope Not Level |
| Breath Volume | | 1.347 | 01:44 | |
| Air Blank | | 0.000 | 01:45 | |
| Air Blank | | 0.000 | 01:47 | |
| Subject Sample | # 2 | | 01:47 | Subject Test Refused |
| Breath Volume | | 0.000 | 01:47 | |
| Air Blank | | 0.000 | 01:47 | |
| Control Test | | 0.078 | 01:48 | |
| Air Blank | | 0.000 | 01:48 | |
| Diagnostics Check | | OK | 01:48 | |

Data Download Date/Time: 08/30/2019 10:41

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VANHOF,J
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: LESLIE

DATE/TIME OF TEST: 08/27/2019 04:35:20

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 04:32 | |
| Air Blank | | 0.000 | 04:32 | |
| Control Test | | 0.080 | 04:33 | |
| Air Blank | | 0.000 | 04:33 | |
| Subject Sample | # 1 | | 04:33 | Subject Test Refused |
| Breath Volume | | 0.000 | 04:33 | |
| Air Blank | | 0.000 | 04:34 | |
| Control Test | | 0.079 | 04:34 | |
| Air Blank | | 0.000 | 04:35 | |
| Diagnostics Check | | OK | 04:35 | |

Data Download Date/Time: 08/30/2019 10:41

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CURRY,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:46

SUBJECT'S LAST NAME: WAY

DATE/TIME OF TEST: 08/27/2019 06:10:21

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 06:07 | |
| Air Blank | | 0.000 | 06:08 | |
| Control Test | | 0.080 | 06:08 | |
| Air Blank | | 0.000 | 06:08 | |
| Subject Sample | # 1 | | 06:08 | Subject Test Refused |
| Breath Volume | | 0.000 | 06:08 | |
| Air Blank | | 0.000 | 06:09 | |
| Control Test | | 0.080 | 06:09 | |
| Air Blank | | 0.000 | 06:10 | |
| Diagnostics Check | | OK | 06:10 | |

Data Download Date/Time: 08/30/2019 10:41

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MCFARLAND,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: COBB

DATE/TIME OF TEST: 08/28/2019 03:00:49

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:53 |
| Air Blank | | 0.000 | 02:54 |
| Control Test | | 0.079 | 02:54 |
| Air Blank | | 0.000 | 02:55 |
| Subject Sample | # 1 | 0.057 | 02:55 |
| Breath Volume | | 1.574 | 02:55 |
| Air Blank | | 0.000 | 02:56 |
| Air Blank | | 0.000 | 02:58 |
| Subject Sample | # 2 | 0.059 | 02:59 |
| Breath Volume | | 2.566 | 02:59 |
| Air Blank | | 0.000 | 02:59 |
| Control Test | | 0.080 | 03:00 |
| Air Blank | | 0.000 | 03:00 |
| Diagnostics Check | | OK | 03:00 |

Data Download Date/Time: 08/30/2019 10:41

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 08/29/2019 02:56:29

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 02:53 | |
| Air Blank | | 0.000 | 02:53 | |
| Control Test | | 0.080 | 02:54 | |
| Air Blank | | 0.000 | 02:54 | |
| Subject Sample | # 1 | | 02:55 | Subject Test Refused |
| Breath Volume | | 0.000 | 02:55 | |
| Air Blank | | 0.000 | 02:55 | |
| Control Test | | 0.080 | 02:55 | |
| Air Blank | | 0.000 | 02:56 | |
| Diagnostics Check | | OK | 02:56 | |

Data Download Date/Time: 08/30/2019 10:41

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATHIS,A,J
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: DELACRUZ,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 07/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:18

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 08/01/2019 18:45:50

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 18:38 |
| Air Blank | | 0.000 | 18:38 |
| Control Test | | 0.078 | 18:39 |
| Air Blank | | 0.000 | 18:39 |
| Subject Sample | # 1 | 0.109 | 18:41 |
| Breath Volume | | 1.128 | 18:41 |
| Air Blank | | 0.000 | 18:41 |
| Air Blank | | 0.000 | 18:43 |
| Subject Sample | # 2 | 0.101 | 18:44 |
| Breath Volume | | 1.867 | 18:44 |
| Air Blank | | 0.000 | 18:44 |
| Control Test | | 0.077 | 18:45 |
| Air Blank | | 0.000 | 18:45 |
| Diagnostics Check | | OK | 18:45 |

Data Download Date/Time: 08/30/2019 08:23

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRES,J,L
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: KNIGHT,G
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 07/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: DUKE

DATE/TIME OF TEST: 08/03/2019 02:58:25

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:50 |
| Air Blank | | 0.000 | 02:50 |
| Control Test | | 0.079 | 02:50 |
| Air Blank | | 0.000 | 02:51 |
| Subject Sample | # 1 | 0.139 | 02:52 |
| Breath Volume | | 1.464 | 02:52 |
| Air Blank | | 0.000 | 02:53 |
| Air Blank | | 0.000 | 02:55 |
| Subject Sample | # 2 | 0.158 | 02:56 |
| Breath Volume | | 1.960 | 02:56 |
| Air Blank | | 0.000 | 02:57 |
| Control Test | | 0.078 | 02:57 |
| Air Blank | | 0.000 | 02:58 |
| Diagnostics Check | | OK | 02:58 |

Data Download Date/Time: 08/30/2019 08:23

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KUTZ,T,E
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: WIGGINS,A
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 07/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: SPRADLEY

DATE/TIME OF TEST: 08/03/2019 04:05:02

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:58 |
| Air Blank | | 0.000 | 03:59 |
| Control Test | | 0.078 | 03:59 |
| Air Blank | | 0.000 | 03:59 |
| Subject Sample | # 1 | 0.161 | 04:00 |
| Breath Volume | | 3.074 | 04:00 |
| Air Blank | | 0.000 | 04:01 |
| Air Blank | | 0.000 | 04:03 |
| Subject Sample | # 2 | 0.158 | 04:03 |
| Breath Volume | | 3.300 | 04:03 |
| Air Blank | | 0.000 | 04:04 |
| Control Test | | 0.078 | 04:04 |
| Air Blank | | 0.000 | 04:04 |
| Diagnostics Check | | OK | 04:05 |

Data Download Date/Time: 08/30/2019 08:23

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,E,E
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 07/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: MENEFEE

DATE/TIME OF TEST: 08/04/2019 01:52:32

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 01:49 | |
| Air Blank | | 0.000 | 01:49 | |
| Control Test | | 0.078 | 01:50 | |
| Air Blank | | 0.000 | 01:50 | |
| Subject Sample | # 1 | | 01:51 | Subject Test Refused |
| Breath Volume | | 0.000 | 01:51 | |
| Air Blank | | 0.000 | 01:51 | |
| Control Test | | 0.079 | 01:51 | |
| Air Blank | | 0.000 | 01:52 | |
| Diagnostics Check | | OK | 01:52 | |

Data Download Date/Time: 08/30/2019 08:23

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KUTZ,T,E
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: KNIGHT,G
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 07/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: COULTER

DATE/TIME OF TEST: 08/04/2019 03:01:30

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 02:58 | |
| Air Blank | | 0.000 | 02:59 | |
| Control Test | | 0.079 | 02:59 | |
| Air Blank | | 0.000 | 02:59 | |
| Subject Sample | # 1 | | 03:00 | Subject Test Refused |
| Breath Volume | | 0.000 | 03:00 | |
| Air Blank | | 0.000 | 03:00 | |
| Control Test | | 0.079 | 03:00 | |
| Air Blank | | 0.000 | 03:01 | |
| Diagnostics Check | | OK | 03:01 | |

Data Download Date/Time: 08/30/2019 08:23

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEDDERLY,G,A
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: DELACRUZ,C
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 07/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:00

SUBJECT'S LAST NAME: HARDEN

DATE/TIME OF TEST: 08/04/2019 07:26:16

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 07:20 | |
| Air Blank | | 0.000 | 07:21 | |
| Control Test | | 0.079 | 07:21 | |
| Air Blank | | 0.000 | 07:22 | |
| Subject Sample | # 1 | | 07:24 | Subject Test Refused |
| Breath Volume | | 0.046 | 07:24 | |
| Air Blank | | 0.000 | 07:25 | |
| Control Test | | 0.079 | 07:25 | |
| Air Blank | | 0.000 | 07:26 | |
| Diagnostics Check | | OK | 07:26 | |

Data Download Date/Time: 08/30/2019 08:23

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBB,CATHERINE,D
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUFFMON,J
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 07/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: WINGATE

DATE/TIME OF TEST: 08/10/2019 02:33:18

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 02:30 | |
| Air Blank | | 0.000 | 02:30 | |
| Control Test | | 0.077 | 02:31 | |
| Air Blank | | 0.000 | 02:31 | |
| Subject Sample | # 1 | | 02:31 | Subject Test Refused |
| Breath Volume | | 0.000 | 02:31 | |
| Air Blank | | 0.000 | 02:32 | |
| Control Test | | 0.078 | 02:32 | |
| Air Blank | | 0.000 | 02:33 | |
| Diagnostics Check | | OK | 02:33 | |

Data Download Date/Time: 08/30/2019 08:23

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATHIS,A,J
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: SANDIFER,R
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 07/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:12

SUBJECT'S LAST NAME: DINKINS

DATE/TIME OF TEST: 08/10/2019 19:39:06

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 19:32 |
| Air Blank | | 0.000 | 19:33 |
| Control Test | | 0.076 | 19:33 |
| Air Blank | | 0.000 | 19:34 |
| Subject Sample | # 1 | 0.178 | 19:34 |
| Breath Volume | | 1.734 | 19:34 |
| Air Blank | | 0.000 | 19:35 |
| Air Blank | | 0.000 | 19:37 |
| Subject Sample | # 2 | 0.173 | 19:37 |
| Breath Volume | | 1.648 | 19:37 |
| Air Blank | | 0.000 | 19:38 |
| Control Test | | 0.077 | 19:38 |
| Air Blank | | 0.000 | 19:39 |
| Diagnostics Check | | OK | 19:39 |

Data Download Date/Time: 08/30/2019 08:23

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBB,CATHERINE,D
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUFFMON,J
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 07/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: MILLS

DATE/TIME OF TEST: 08/11/2019 01:02:21

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:55 |
| Air Blank | | 0.000 | 00:55 |
| Control Test | | 0.078 | 00:56 |
| Air Blank | | 0.000 | 00:56 |
| Subject Sample | # 1 | 0.103 | 00:57 |
| Breath Volume | | 1.542 | 00:57 |
| Air Blank | | 0.000 | 00:58 |
| Air Blank | | 0.000 | 01:00 |
| Subject Sample | # 2 | 0.103 | 01:00 |
| Breath Volume | | 1.671 | 01:00 |
| Air Blank | | 0.000 | 01:01 |
| Control Test | | 0.077 | 01:01 |
| Air Blank | | 0.000 | 01:02 |
| Diagnostics Check | | OK | 01:02 |

Data Download Date/Time: 08/30/2019 08:23

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBB,CATHERINE,D
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUFFMON,J
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 07/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: MILLS

DATE/TIME OF TEST: 08/11/2019 01:31:47

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:25 |
| Air Blank | | 0.000 | 01:25 |
| Control Test | | 0.078 | 01:25 |
| Air Blank | | 0.000 | 01:26 |
| Subject Sample | # 1 | 0.088 | 01:27 |
| Breath Volume | | 1.765 | 01:27 |
| Air Blank | | 0.000 | 01:27 |
| Air Blank | | 0.000 | 01:29 |
| Subject Sample | # 2 | 0.092 | 01:30 |
| Breath Volume | | 1.382 | 01:30 |
| Air Blank | | 0.000 | 01:30 |
| Control Test | | 0.077 | 01:31 |
| Air Blank | | 0.000 | 01:31 |
| Diagnostics Check | | OK | 01:31 |

Data Download Date/Time: 08/30/2019 08:23

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATHIS,A,J
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUFFMON,J
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 07/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: DEATON

DATE/TIME OF TEST: 08/11/2019 04:15:09

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:08 |
| Air Blank | | 0.000 | 04:08 |
| Control Test | | 0.078 | 04:09 |
| Air Blank | | 0.000 | 04:09 |
| Subject Sample | # 1 | 0.109 | 04:10 |
| Breath Volume | | 1.421 | 04:10 |
| Air Blank | | 0.000 | 04:11 |
| Air Blank | | 0.000 | 04:13 |
| Subject Sample | # 2 | 0.097 | 04:13 |
| Breath Volume | | 1.160 | 04:13 |
| Air Blank | | 0.000 | 04:14 |
| Control Test | | 0.078 | 04:14 |
| Air Blank | | 0.000 | 04:15 |
| Diagnostics Check | | OK | 04:15 |

Data Download Date/Time: 08/30/2019 08:23

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBB,CATHERINE,D
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: DELACRUZ,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 07/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: SWEENEY

DATE/TIME OF TEST: 08/11/2019 23:01:06

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 22:58 | |
| Air Blank | | 0.000 | 22:58 | |
| Control Test | | 0.078 | 22:58 | |
| Air Blank | | 0.000 | 22:59 | |
| Subject Sample | # 1 | | 22:59 | Subject Test Refused |
| Breath Volume | | 0.000 | 22:59 | |
| Air Blank | | 0.000 | 23:00 | |
| Control Test | | 0.079 | 23:00 | |
| Air Blank | | 0.000 | 23:01 | |
| Diagnostics Check | | OK | 23:01 | |

Data Download Date/Time: 08/30/2019 08:23

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATHIS,A,J
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: SANDIFER,R
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 07/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: RIVERA MELENDEZ

DATE/TIME OF TEST: 08/11/2019 23:39:53

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:33 |
| Air Blank | | 0.000 | 23:34 |
| Control Test | | 0.078 | 23:34 |
| Air Blank | | 0.000 | 23:35 |
| Subject Sample | # 1 | 0.247 | 23:35 |
| Breath Volume | | 1.625 | 23:35 |
| Air Blank | | 0.000 | 23:36 |
| Air Blank | | 0.000 | 23:37 |
| Subject Sample | # 2 | 0.236 | 23:38 |
| Breath Volume | | 1.203 | 23:38 |
| Air Blank | | 0.000 | 23:38 |
| Control Test | | 0.078 | 23:39 |
| Air Blank | | 0.000 | 23:39 |
| Diagnostics Check | | OK | 23:39 |

Data Download Date/Time: 08/30/2019 08:23

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATHIS,A,J
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUFFMON,J
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 07/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: VAUGHN

DATE/TIME OF TEST: 08/15/2019 22:23:32

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 22:20 | |
| Air Blank | | 0.000 | 22:21 | |
| Control Test | | 0.076 | 22:21 | |
| Air Blank | | 0.000 | 22:21 | |
| Subject Sample | # 1 | | 22:22 | Subject Test Refused |
| Breath Volume | | 0.000 | 22:22 | |
| Air Blank | | 0.000 | 22:22 | |
| Control Test | | 0.077 | 22:22 | |
| Air Blank | | 0.000 | 22:23 | |
| Diagnostics Check | | OK | 22:23 | |

Data Download Date/Time: 08/30/2019 08:23

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBB,CATHERINE,D
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUFFMON,J
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 07/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: ANSON

DATE/TIME OF TEST: 08/16/2019 04:39:41

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 04:28 | |
| Air Blank | | 0.000 | 04:28 | |
| Control Test | | 0.079 | 04:28 | |
| Air Blank | | 0.000 | 04:29 | |
| Subject Sample | # 1 | 0.120 | 04:32 | Volume Not Met |
| Breath Volume | | 0.222 | 04:32 | |
| Air Blank | | 0.000 | 04:33 | |
| Air Blank | | 0.000 | 04:35 | |
| Subject Sample | # 2 | 0.104 | 04:38 | Volume Not Met |
| Breath Volume | | 0.015 | 04:38 | |
| Air Blank | | 0.000 | 04:38 | |
| Control Test | | 0.078 | 04:39 | |
| Air Blank | | 0.000 | 04:39 | |
| Diagnostics Check | | OK | 04:39 | |

Data Download Date/Time: 08/30/2019 08:23

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRES,J,L
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 07/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: BULLARD

DATE/TIME OF TEST: 08/17/2019 02:45:03

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 02:42 | |
| Air Blank | | 0.000 | 02:42 | |
| Control Test | | 0.079 | 02:42 | |
| Air Blank | | 0.000 | 02:43 | |
| Subject Sample | # 1 | | 02:43 | Subject Test Refused |
| Breath Volume | | 0.000 | 02:43 | |
| Air Blank | | 0.000 | 02:44 | |
| Control Test | | 0.079 | 02:44 | |
| Air Blank | | 0.000 | 02:44 | |
| Diagnostics Check | | OK | 02:45 | |

Data Download Date/Time: 08/30/2019 08:23

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KUTZ,T,E
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 07/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 08/18/2019 02:24:40

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 02:21 | |
| Air Blank | | 0.000 | 02:22 | |
| Control Test | | 0.078 | 02:22 | |
| Air Blank | | 0.000 | 02:23 | |
| Subject Sample | # 1 | | 02:23 | Subject Test Refused |
| Breath Volume | | 0.000 | 02:23 | |
| Air Blank | | 0.000 | 02:23 | |
| Control Test | | 0.079 | 02:24 | |
| Air Blank | | 0.000 | 02:24 | |
| Diagnostics Check | | OK | 02:24 | |

Data Download Date/Time: 08/30/2019 08:23

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCRACKEN,J,D
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 07/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: MCGOWEN

DATE/TIME OF TEST: 08/18/2019 23:14:13

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 23:03 | |
| Air Blank | | 0.000 | 23:03 | |
| Control Test | | 0.078 | 23:04 | |
| Air Blank | | 0.000 | 23:04 | |
| Subject Sample | # 1 | | 23:07 | No Sample Provided |
| Breath Volume | | 0.000 | 23:07 | |
| Air Blank | | 0.000 | 23:08 | |
| Air Blank | | 0.000 | 23:10 | |
| Subject Sample | # 2 | | 23:12 | Subject Test Refused |
| Breath Volume | | 0.000 | 23:12 | |
| Air Blank | | 0.000 | 23:13 | |
| Control Test | | 0.078 | 23:13 | |
| Air Blank | | 0.000 | 23:14 | |
| Diagnostics Check | | OK | 23:14 | |

Data Download Date/Time: 08/30/2019 08:23

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,E,E
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: KNIGHT,G
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 07/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 08/19/2019 04:14:08

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 04:10 | |
| Air Blank | | 0.000 | 04:11 | |
| Control Test | | 0.079 | 04:11 | |
| Air Blank | | 0.000 | 04:12 | |
| Subject Sample | # 1 | | 04:12 | Subject Test Refused |
| Breath Volume | | 0.000 | 04:12 | |
| Air Blank | | 0.000 | 04:13 | |
| Control Test | | 0.078 | 04:13 | |
| Air Blank | | 0.000 | 04:14 | |
| Diagnostics Check | | OK | 04:14 | |

Data Download Date/Time: 08/30/2019 08:23

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATHIS,A,J
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 07/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: RADTKE

DATE/TIME OF TEST: 08/19/2019 23:53:41

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 23:50 | |
| Air Blank | | 0.000 | 23:51 | |
| Control Test | | 0.078 | 23:51 | |
| Air Blank | | 0.000 | 23:52 | |
| Subject Sample | # 1 | | 23:52 | Subject Test Refused |
| Breath Volume | | 0.000 | 23:52 | |
| Air Blank | | 0.000 | 23:52 | |
| Control Test | | 0.078 | 23:53 | |
| Air Blank | | 0.000 | 23:53 | |
| Diagnostics Check | | OK | 23:53 | |

Data Download Date/Time: 08/30/2019 08:23

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIGGINBOTHAM,B,K
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: REYNOLDS,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 07/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: KLIX

DATE/TIME OF TEST: 08/20/2019 21:28:57

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 21:25 | |
| Air Blank | | 0.000 | 21:26 | |
| Control Test | | 0.079 | 21:26 | |
| Air Blank | | 0.000 | 21:27 | |
| Subject Sample | # 1 | | 21:27 | Subject Test Refused |
| Breath Volume | | 0.000 | 21:27 | |
| Air Blank | | 0.000 | 21:28 | |
| Control Test | | 0.078 | 21:28 | |
| Air Blank | | 0.000 | 21:28 | |
| Diagnostics Check | | OK | 21:28 | |

Data Download Date/Time: 08/30/2019 08:23

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATHIS,A,J
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 07/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: FORD

DATE/TIME OF TEST: 08/21/2019 01:50:27

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:43 |
| Air Blank | | 0.000 | 01:44 |
| Control Test | | 0.078 | 01:44 |
| Air Blank | | 0.000 | 01:44 |
| Subject Sample | # 1 | 0.133 | 01:45 |
| Breath Volume | | 1.308 | 01:45 |
| Air Blank | | 0.000 | 01:46 |
| Air Blank | | 0.000 | 01:48 |
| Subject Sample | # 2 | 0.131 | 01:48 |
| Breath Volume | | 1.203 | 01:48 |
| Air Blank | | 0.000 | 01:49 |
| Control Test | | 0.078 | 01:49 |
| Air Blank | | 0.000 | 01:50 |
| Diagnostics Check | | OK | 01:50 |

Data Download Date/Time: 08/30/2019 08:23

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATHIS,A,J
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: KNIGHT,G
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 07/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: STEINER

DATE/TIME OF TEST: 08/21/2019 04:51:23

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:44 |
| Air Blank | | 0.000 | 04:45 |
| Control Test | | 0.079 | 04:45 |
| Air Blank | | 0.000 | 04:46 |
| Subject Sample | # 1 | 0.119 | 04:46 |
| Breath Volume | | 1.585 | 04:46 |
| Air Blank | | 0.000 | 04:47 |
| Air Blank | | 0.000 | 04:49 |
| Subject Sample | # 2 | 0.113 | 04:49 |
| Breath Volume | | 1.417 | 04:49 |
| Air Blank | | 0.000 | 04:50 |
| Control Test | | 0.077 | 04:50 |
| Air Blank | | 0.000 | 04:51 |
| Diagnostics Check | | OK | 04:51 |

Data Download Date/Time: 08/30/2019 08:23

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIGGINBOTHAM,B,K
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: KNIGHT,B
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 07/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: DEGAETANO

DATE/TIME OF TEST: 08/21/2019 05:26:52

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 05:20 |
| Air Blank | | 0.000 | 05:20 |
| Control Test | | 0.077 | 05:21 |
| Air Blank | | 0.000 | 05:21 |
| Subject Sample | # 1 | 0.229 | 05:22 |
| Breath Volume | | 1.554 | 05:22 |
| Air Blank | | 0.000 | 05:22 |
| Air Blank | | 0.000 | 05:24 |
| Subject Sample | # 2 | 0.228 | 05:25 |
| Breath Volume | | 1.769 | 05:25 |
| Air Blank | | 0.000 | 05:25 |
| Control Test | | 0.079 | 05:26 |
| Air Blank | | 0.000 | 05:26 |
| Diagnostics Check | | OK | 05:26 |

Data Download Date/Time: 08/30/2019 08:23

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRES,J,L
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: JACKSON,R
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 07/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: GODWIN

DATE/TIME OF TEST: 08/22/2019 02:27:44

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:21 |
| Air Blank | | 0.000 | 02:22 |
| Control Test | | 0.078 | 02:22 |
| Air Blank | | 0.000 | 02:23 |
| Subject Sample | # 1 | 0.000 | 02:23 |
| Breath Volume | | 2.804 | 02:23 |
| Air Blank | | 0.000 | 02:23 |
| Air Blank | | 0.000 | 02:26 |
| Subject Sample | # 2 | 0.000 | 02:26 |
| Breath Volume | | 2.542 | 02:26 |
| Air Blank | | 0.000 | 02:26 |
| Control Test | | 0.078 | 02:27 |
| Air Blank | | 0.000 | 02:27 |
| Diagnostics Check | | OK | 02:27 |

Data Download Date/Time: 08/30/2019 08:23

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,E,E
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 07/21/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: BOLT

DATE/TIME OF TEST: 08/23/2019 01:42:10

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:35 |
| Air Blank | | 0.000 | 01:36 |
| Control Test | | 0.079 | 01:36 |
| Air Blank | | 0.000 | 01:36 |
| Subject Sample | # 1 | 0.025 | 01:37 |
| Breath Volume | | 3.183 | 01:37 |
| Air Blank | | 0.000 | 01:38 |
| Air Blank | | 0.000 | 01:40 |
| Subject Sample | # 2 | 0.024 | 01:40 |
| Breath Volume | | 3.292 | 01:40 |
| Air Blank | | 0.000 | 01:41 |
| Control Test | | 0.078 | 01:41 |
| Air Blank | | 0.000 | 01:42 |
| Diagnostics Check | | OK | 01:42 |

Data Download Date/Time: 08/30/2019 08:23

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,E,E
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: WOODS,W
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 07/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: CARPENTER

DATE/TIME OF TEST: 08/23/2019 02:24:27

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:16 |
| Air Blank | | 0.000 | 02:16 |
| Control Test | | 0.077 | 02:16 |
| Air Blank | | 0.000 | 02:17 |
| Subject Sample | # 1 | 0.197 | 02:18 |
| Breath Volume | | 1.250 | 02:18 |
| Air Blank | | 0.000 | 02:18 |
| Air Blank | | 0.000 | 02:20 |
| Subject Sample | # 2 | 0.208 | 02:22 |
| Breath Volume | | 2.480 | 02:22 |
| Air Blank | | 0.000 | 02:23 |
| Control Test | | 0.076 | 02:23 |
| Air Blank | | 0.000 | 02:24 |
| Diagnostics Check | | OK | 02:24 |

Data Download Date/Time: 08/30/2019 08:23

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBB,CATHERINE,D
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 07/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: NORRIS

DATE/TIME OF TEST: 08/25/2019 01:34:08

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:27 |
| Air Blank | | 0.000 | 01:28 |
| Control Test | | 0.078 | 01:28 |
| Air Blank | | 0.000 | 01:29 |
| Subject Sample | # 1 | 0.179 | 01:29 |
| Breath Volume | | 1.632 | 01:29 |
| Air Blank | | 0.000 | 01:30 |
| Air Blank | | 0.000 | 01:32 |
| Subject Sample | # 2 | 0.181 | 01:32 |
| Breath Volume | | 1.683 | 01:32 |
| Air Blank | | 0.000 | 01:33 |
| Control Test | | 0.077 | 01:33 |
| Air Blank | | 0.000 | 01:34 |
| Diagnostics Check | | OK | 01:34 |

Data Download Date/Time: 08/30/2019 08:23

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIGGINBOTHAM,B,K
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: BALLARD,J
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 07/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: HAYES

DATE/TIME OF TEST: 08/25/2019 02:21:29

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:10 |
| Air Blank | | 0.000 | 02:10 |
| Control Test | | 0.077 | 02:10 |
| Air Blank | | 0.000 | 02:11 |
| Subject Sample | # 1 | 0.238 | 02:12 |
| Breath Volume | | 1.714 | 02:12 |
| Air Blank | | 0.000 | 02:13 |
| Air Blank | | 0.000 | 02:15 |
| Subject Sample | # 2 | 0.259 | 02:15 |
| Breath Volume | | 1.218 | 02:15 |
| Air Blank | | 0.000 | 02:16 |
| Air Blank | | 0.000 | 02:18 |
| Subject Sample | # 3 | 0.219 | 02:19 |
| Breath Volume | | 1.406 | 02:19 |
| Air Blank | | 0.000 | 02:20 |
| Control Test | | 0.077 | 02:20 |
| Air Blank | | 0.000 | 02:21 |
| Diagnostics Check | | OK | 02:21 |

Data Download Date/Time: 08/30/2019 08:23

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBB,CATHERINE,D
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUFFMON,J
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 07/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: SMALLWOOD

DATE/TIME OF TEST: 08/25/2019 02:47:29

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 02:44 | |
| Air Blank | | 0.000 | 02:45 | |
| Control Test | | 0.078 | 02:45 | |
| Air Blank | | 0.000 | 02:45 | |
| Subject Sample | # 1 | | 02:46 | Subject Test Refused |
| Breath Volume | | 0.000 | 02:46 | |
| Air Blank | | 0.000 | 02:46 | |
| Control Test | | 0.079 | 02:46 | |
| Air Blank | | 0.000 | 02:47 | |
| Diagnostics Check | | OK | 02:47 | |

Data Download Date/Time: 08/30/2019 08:23

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCRACKEN,J,D
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 07/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 08/28/2019 00:51:23

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 00:48 | |
| Air Blank | | 0.000 | 00:48 | |
| Control Test | | 0.079 | 00:49 | |
| Air Blank | | 0.000 | 00:49 | |
| Subject Sample | # 1 | | 00:50 | Subject Test Refused |
| Breath Volume | | 0.000 | 00:50 | |
| Air Blank | | 0.000 | 00:50 | |
| Control Test | | 0.080 | 00:50 | |
| Air Blank | | 0.000 | 00:51 | |
| Diagnostics Check | | OK | 00:51 | |

Data Download Date/Time: 08/30/2019 08:23

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KUTZ,T,E
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: JACKSON,R
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 07/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: BUNTING

DATE/TIME OF TEST: 08/28/2019 01:48:40

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:42 |
| Air Blank | | 0.000 | 01:42 |
| Control Test | | 0.079 | 01:42 |
| Air Blank | | 0.000 | 01:43 |
| Subject Sample | # 1 | 0.180 | 01:44 |
| Breath Volume | | 1.960 | 01:44 |
| Air Blank | | 0.000 | 01:44 |
| Air Blank | | 0.000 | 01:46 |
| Subject Sample | # 2 | 0.176 | 01:47 |
| Breath Volume | | 2.160 | 01:47 |
| Air Blank | | 0.000 | 01:47 |
| Control Test | | 0.077 | 01:48 |
| Air Blank | | 0.000 | 01:48 |
| Diagnostics Check | | OK | 01:48 |

Data Download Date/Time: 08/30/2019 08:23

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATHIS,A,J
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 07/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: SWARTZ

DATE/TIME OF TEST: 08/29/2019 02:57:49

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:50 |
| Air Blank | | 0.000 | 02:51 |
| Control Test | | 0.078 | 02:51 |
| Air Blank | | 0.000 | 02:52 |
| Subject Sample | # 1 | 0.251 | 02:52 |
| Breath Volume | | 1.265 | 02:52 |
| Air Blank | | 0.000 | 02:53 |
| Air Blank | | 0.000 | 02:55 |
| Subject Sample | # 2 | 0.260 | 02:56 |
| Breath Volume | | 1.593 | 02:56 |
| Air Blank | | 0.000 | 02:56 |
| Control Test | | 0.077 | 02:57 |
| Air Blank | | 0.000 | 02:57 |
| Diagnostics Check | | OK | 02:57 |

Data Download Date/Time: 08/30/2019 08:23

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIGGINBOTHAM,B,K
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: SANDIFER,R
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 07/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:02

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 08/29/2019 20:28:55

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 20:22 |
| Air Blank | | 0.000 | 20:22 |
| Control Test | | 0.078 | 20:23 |
| Air Blank | | 0.000 | 20:23 |
| Subject Sample | # 1 | 0.123 | 20:24 |
| Breath Volume | | 1.457 | 20:24 |
| Air Blank | | 0.000 | 20:25 |
| Air Blank | | 0.000 | 20:27 |
| Subject Sample | # 2 | 0.128 | 20:27 |
| Breath Volume | | 2.128 | 20:27 |
| Air Blank | | 0.000 | 20:27 |
| Control Test | | 0.078 | 20:28 |
| Air Blank | | 0.000 | 20:28 |
| Diagnostics Check | | OK | 20:28 |

Data Download Date/Time: 08/30/2019 08:24

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001285

INSTRUMENT REGISTERED TO: GILCHRIST COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANUS,GLEN,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: GANUS,GLEN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1048227
DRY GAS EXPIRY DATE: 11/15/2021
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 08/18/2019 03:52:17

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:44 |
| Air Blank | | 0.000 | 03:45 |
| Control Test | | 0.080 | 03:45 |
| Air Blank | | 0.000 | 03:46 |
| Subject Sample | # 1 | 0.120 | 03:46 |
| Breath Volume | | 1.347 | 03:46 |
| Air Blank | | 0.000 | 03:47 |
| Air Blank | | 0.000 | 03:49 |
| Subject Sample | # 2 | 0.115 | 03:50 |
| Breath Volume | | 1.332 | 03:50 |
| Air Blank | | 0.000 | 03:51 |
| Control Test | | 0.079 | 03:51 |
| Air Blank | | 0.000 | 03:52 |
| Diagnostics Check | | OK | 03:52 |

Data Download Date/Time: 08/29/2019 02:30

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

INSTRUMENT REGISTERED TO: BAKER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILLIARD,KARI,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HILLIARD,KARI
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 07/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 08/23/2019 03:34:26

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:27 |
| Air Blank | | 0.000 | 03:28 |
| Control Test | | 0.080 | 03:28 |
| Air Blank | | 0.000 | 03:29 |
| Subject Sample | # 1 | 0.000 | 03:30 |
| Breath Volume | | 1.742 | 03:30 |
| Air Blank | | 0.000 | 03:30 |
| Air Blank | | 0.000 | 03:32 |
| Subject Sample | # 2 | 0.000 | 03:33 |
| Breath Volume | | 3.332 | 03:33 |
| Air Blank | | 0.000 | 03:33 |
| Control Test | | 0.079 | 03:33 |
| Air Blank | | 0.000 | 03:34 |
| Diagnostics Check | | OK | 03:34 |

Data Download Date/Time: 08/29/2019 20:44

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

INSTRUMENT REGISTERED TO: BAKER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILLIARD,KARI,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HILLIARD,KARI
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 07/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: REED

DATE/TIME OF TEST: 08/25/2019 01:51:08

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 01:35 | |
| Air Blank | | 0.000 | 01:36 | |
| Control Test | | 0.080 | 01:36 | |
| Air Blank | | 0.000 | 01:37 | |
| Subject Sample | # 1 | 0.157 | 01:39 | |
| Breath Volume | | 1.367 | 01:39 | |
| Air Blank | | 0.000 | 01:40 | |
| Air Blank | | 0.000 | 01:42 | |
| Subject Sample | # 2 | 0.143 | 01:45 | Volume Not Met |
| Breath Volume | | 0.277 | 01:45 | |
| Air Blank | | 0.000 | 01:46 | |
| Air Blank | | 0.000 | 01:47 | |
| Subject Sample | # 3 | 0.154 | 01:49 | |
| Breath Volume | | 1.507 | 01:49 | |
| Air Blank | | 0.000 | 01:50 | |
| Control Test | | 0.077 | 01:50 | |
| Air Blank | | 0.000 | 01:51 | |
| Diagnostics Check | | OK | 01:51 | |

Data Download Date/Time: 08/29/2019 20:44

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001288

INSTRUMENT REGISTERED TO: UNION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAEYAERT,JEFF,H
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: NAEYAERT,JEFF
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 823252
DRY GAS EXPIRY DATE: 08/07/2020
VIOLATION CODE: DUI
ARREST TIME: 15:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:35

SUBJECT'S LAST NAME: GEORGE

DATE/TIME OF TEST: 08/29/2019 16:23:09

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 16:16 |
| Air Blank | | 0.000 | 16:16 |
| Control Test | | 0.080 | 16:16 |
| Air Blank | | 0.000 | 16:17 |
| Subject Sample | # 1 | 0.138 | 16:17 |
| Breath Volume | | 1.144 | 16:17 |
| Air Blank | | 0.000 | 16:18 |
| Air Blank | | 0.000 | 16:20 |
| Subject Sample | # 2 | 0.136 | 16:21 |
| Breath Volume | | 2.085 | 16:21 |
| Air Blank | | 0.000 | 16:22 |
| Control Test | | 0.078 | 16:22 |
| Air Blank | | 0.000 | 16:23 |
| Diagnostics Check | | OK | 16:23 |

Data Download Date/Time: 09/03/2019 09:29

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HULLENDER,ROBERT,S
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: FLYNN,HALEY
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 07/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: COBB

DATE/TIME OF TEST: 08/06/2019 02:43:33

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:30 |
| Air Blank | | 0.000 | 02:31 |
| Control Test | | 0.078 | 02:31 |
| Air Blank | | 0.000 | 02:32 |
| Subject Sample | # 1 | 0.174 | 02:33 |
| Breath Volume | | 3.109 | 02:33 |
| Air Blank | | 0.000 | 02:33 |
| Air Blank | | 0.000 | 02:35 |
| Subject Sample | # 2 | 0.150 | 02:36 |
| Breath Volume | | 1.148 | 02:36 |
| Air Blank | | 0.000 | 02:36 |
| Air Blank | | 0.000 | 02:38 |
| Subject Sample | # 3 | 0.153 | 02:41 |
| Breath Volume | | 1.558 | 02:41 |
| Air Blank | | 0.000 | 02:42 |
| Control Test | | 0.077 | 02:42 |
| Air Blank | | 0.000 | 02:43 |
| Diagnostics Check | | OK | 02:43 |

Data Download Date/Time: 09/05/2019 07:07

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HULLENDER,ROBERT,S
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: ROONEY,MICHAEL
ARRESTING OFFICER AGENCY: Starke PD
LAST AGENCY INSPECTION: 07/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: FRANCIS

DATE/TIME OF TEST: 08/15/2019 03:01:07

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:54 |
| Air Blank | | 0.000 | 02:55 |
| Control Test | | 0.078 | 02:55 |
| Air Blank | | 0.000 | 02:56 |
| Subject Sample | # 1 | 0.196 | 02:56 |
| Breath Volume | | 2.972 | 02:56 |
| Air Blank | | 0.000 | 02:57 |
| Air Blank | | 0.000 | 02:58 |
| Subject Sample | # 2 | 0.199 | 02:59 |
| Breath Volume | | 3.562 | 02:59 |
| Air Blank | | 0.000 | 03:00 |
| Control Test | | 0.076 | 03:00 |
| Air Blank | | 0.000 | 03:01 |
| Diagnostics Check | | OK | 03:01 |

Data Download Date/Time: 09/05/2019 07:07

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HULLENDER,ROBERT,S
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: FLYNN,HALEY
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 07/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 18:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:52

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 08/24/2019 19:15:37

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 19:12 | |
| Air Blank | | 0.000 | 19:12 | |
| Control Test | | 0.077 | 19:13 | |
| Air Blank | | 0.000 | 19:13 | |
| Subject Sample | # 1 | | 19:14 | Subject Test Refused |
| Breath Volume | | 0.000 | 19:14 | |
| Air Blank | | 0.000 | 19:14 | |
| Control Test | | 0.077 | 19:15 | |
| Air Blank | | 0.000 | 19:15 | |
| Diagnostics Check | | OK | 19:15 | |

Data Download Date/Time: 09/05/2019 07:07

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HULLENDER,ROBERT,S
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: TOMPKINS,DANIEL
ARRESTING OFFICER AGENCY: Lawtey PD
LAST AGENCY INSPECTION: 07/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: DUDLEY

DATE/TIME OF TEST: 08/25/2019 22:57:47

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 22:54 | |
| Air Blank | | 0.000 | 22:55 | |
| Control Test | | 0.078 | 22:55 | |
| Air Blank | | 0.000 | 22:56 | |
| Subject Sample | # 1 | | 22:56 | Subject Test Refused |
| Breath Volume | | 0.000 | 22:56 | |
| Air Blank | | 0.000 | 22:56 | |
| Control Test | | 0.078 | 22:57 | |
| Air Blank | | 0.000 | 22:57 | |
| Diagnostics Check | | OK | 22:57 | |

Data Download Date/Time: 09/05/2019 07:07

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HULLENDER,ROBERT,S
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: HULLENDER,ROBERT
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 08/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 08/28/2019 23:09:20

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 23:05 | |
| Air Blank | | 0.000 | 23:05 | |
| Control Test | | 0.077 | 23:05 | |
| Air Blank | | 0.000 | 23:06 | |
| Subject Sample | # 1 | | 23:07 | Subject Test Refused |
| Breath Volume | | 0.085 | 23:07 | |
| Air Blank | | 0.000 | 23:08 | |
| Control Test | | 0.078 | 23:08 | |
| Air Blank | | 0.000 | 23:09 | |
| Diagnostics Check | | OK | 23:09 | |

Data Download Date/Time: 09/13/2019 07:13

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEDENBAUGH,JUSTIN,H
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BEDENBAUGH,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805701
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 13:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:02

SUBJECT'S LAST NAME: JOHN

DATE/TIME OF TEST: 08/01/2019 14:49:09

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 14:46 | |
| Air Blank | | 0.000 | 14:46 | |
| Control Test | | 0.078 | 14:47 | |
| Air Blank | | 0.000 | 14:47 | |
| Subject Sample | # 1 | | 14:47 | Subject Test Refused |
| Breath Volume | | 0.000 | 14:47 | |
| Air Blank | | 0.000 | 14:48 | |
| Control Test | | 0.078 | 14:48 | |
| Air Blank | | 0.000 | 14:49 | |
| Diagnostics Check | | OK | 14:49 | |

Data Download Date/Time: 09/04/2019 17:15

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASH,EARL,C
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MASH,EARL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805701
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 11:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:40

SUBJECT'S LAST NAME: MERRILL

DATE/TIME OF TEST: 08/09/2019 13:08:30

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 13:01 |
| Air Blank | | 0.000 | 13:02 |
| Control Test | | 0.078 | 13:02 |
| Air Blank | | 0.000 | 13:02 |
| Subject Sample | # 1 | 0.000 | 13:04 |
| Breath Volume | | 3.523 | 13:04 |
| Air Blank | | 0.000 | 13:04 |
| Air Blank | | 0.000 | 13:06 |
| Subject Sample | # 2 | 0.000 | 13:07 |
| Breath Volume | | 3.914 | 13:07 |
| Air Blank | | 0.000 | 13:07 |
| Control Test | | 0.078 | 13:07 |
| Air Blank | | 0.000 | 13:08 |
| Diagnostics Check | | OK | 13:08 |

Data Download Date/Time: 09/04/2019 17:15

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRYDEN,JOHN,E
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: DRYDEN,JOHN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805701
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: HATCHER

DATE/TIME OF TEST: 08/25/2019 01:27:20

| | | | |
|---|---|---|---|
| Diagnostics Check | OK | 01:25 | |
| Air Blank | 0.000 | 01:26 | |
| Control Test | 0.079 | 01:26 | |
| Air Blank | | 01:26 | RFI Detect |
| Air Blank | 0.000 | 01:27 | |

Data Download Date/Time: 09/04/2019 17:15

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRYDEN,JOHN,E
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: DRYDEN,JOHN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805701
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: HATCHER

DATE/TIME OF TEST: 08/25/2019 01:40:18

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:33 |
| Air Blank | | 0.000 | 01:34 |
| Control Test | | 0.078 | 01:34 |
| Air Blank | | 0.000 | 01:35 |
| Subject Sample | # 1 | 0.000 | 01:35 |
| Breath Volume | | 4.378 | 01:35 |
| Air Blank | | 0.000 | 01:36 |
| Air Blank | | 0.000 | 01:38 |
| Subject Sample | # 2 | 0.000 | 01:38 |
| Breath Volume | | 4.425 | 01:38 |
| Air Blank | | 0.000 | 01:39 |
| Control Test | | 0.078 | 01:39 |
| Air Blank | | 0.000 | 01:40 |
| Diagnostics Check | | OK | 01:40 |

Data Download Date/Time: 09/04/2019 17:15

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLK,AUSTIN,B
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: NA,NA
ARRESTING OFFICER AGENCY: Levy County SO
LAST AGENCY INSPECTION: 07/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805701
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: COURT ORD
ARREST TIME: 17:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 08/26/2019 19:37:15

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 19:30 |
| Air Blank | | 0.000 | 19:31 |
| Control Test | | 0.079 | 19:31 |
| Air Blank | | 0.000 | 19:32 |
| Subject Sample | # 1 | 0.008 | 19:33 |
| Breath Volume | | 2.648 | 19:33 |
| Air Blank | | 0.000 | 19:33 |
| Air Blank | | 0.000 | 19:35 |
| Subject Sample | # 2 | 0.000 | 19:35 |
| Breath Volume | | 1.621 | 19:35 |
| Air Blank | | 0.000 | 19:36 |
| Control Test | | 0.077 | 19:36 |
| Air Blank | | 0.000 | 19:37 |
| Diagnostics Check | | OK | 19:37 |

Data Download Date/Time: 09/04/2019 17:15

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLK,AUSTIN,B
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: NA,NA
ARRESTING OFFICER AGENCY: Levy County SO
LAST AGENCY INSPECTION: 07/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805701
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: COURT ORD
ARREST TIME: 19:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:21

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 08/26/2019 20:03:22

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 19:56 |
| Air Blank | | 0.000 | 19:57 |
| Control Test | | 0.078 | 19:57 |
| Air Blank | | 0.000 | 19:58 |
| Subject Sample | # 1 | 0.000 | 19:58 |
| Breath Volume | | 2.828 | 19:58 |
| Air Blank | | 0.000 | 19:59 |
| Air Blank | | 0.000 | 20:01 |
| Subject Sample | # 2 | 0.000 | 20:02 |
| Breath Volume | | 3.140 | 20:02 |
| Air Blank | | 0.000 | 20:02 |
| Control Test | | 0.078 | 20:02 |
| Air Blank | | 0.000 | 20:03 |
| Diagnostics Check | | OK | 20:03 |

Data Download Date/Time: 09/04/2019 17:15

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ECKER,DEREK,C
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: CHERAMIE,MICAH
ARRESTING OFFICER AGENCY: Chiefland PD
LAST AGENCY INSPECTION: 07/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805701
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: MAILLARD

DATE/TIME OF TEST: 08/29/2019 02:15:48

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 02:13 | |
| Air Blank | | 0.000 | 02:13 | |
| Control Test | | 0.078 | 02:13 | |
| Air Blank | | 0.000 | 02:14 | |
| Subject Sample | # 1 | | 02:14 | Subject Test Refused |
| Breath Volume | | 0.000 | 02:14 | |
| Air Blank | | 0.000 | 02:14 | |
| Control Test | | 0.078 | 02:15 | |
| Air Blank | | 0.000 | 02:15 | |
| Diagnostics Check | | OK | 02:15 | |

Data Download Date/Time: 09/04/2019 17:15

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANUS,GLEN,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: GANUS,GLEN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805701
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 08/31/2019 04:28:45

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:20 |
| Air Blank | | 0.000 | 04:20 |
| Control Test | | 0.079 | 04:20 |
| Air Blank | | 0.000 | 04:21 |
| Subject Sample | # 1 | 0.222 | 04:23 |
| Breath Volume | | 1.207 | 04:23 |
| Air Blank | | 0.000 | 04:24 |
| Air Blank | | 0.000 | 04:26 |
| Subject Sample | # 2 | 0.235 | 04:27 |
| Breath Volume | | 1.542 | 04:27 |
| Air Blank | | 0.000 | 04:27 |
| Control Test | | 0.077 | 04:28 |
| Air Blank | | 0.000 | 04:28 |
| Diagnostics Check | | OK | 04:28 |

Data Download Date/Time: 09/04/2019 17:15

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001293

INSTRUMENT REGISTERED TO: WILLISTON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIRON,KENNY,R
BREATH TEST OPERATOR AGENCY: Williston PD
ARRESTING OFFICER: ROSA,C
ARRESTING OFFICER AGENCY: Williston PD
LAST AGENCY INSPECTION: 07/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 937240
DRY GAS EXPIRY DATE: 04/07/2021
VIOLATION CODE: DUI
ARREST TIME: 05:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:56

SUBJECT'S LAST NAME: GIRON

DATE/TIME OF TEST: 08/24/2019 06:26:48

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 06:23 | |
| Air Blank | | 0.000 | 06:24 | |
| Control Test | | 0.079 | 06:24 | |
| Air Blank | | 0.000 | 06:25 | |
| Subject Sample | # 1 | | 06:25 | Subject Test Refused |
| Breath Volume | | 0.000 | 06:25 | |
| Air Blank | | 0.000 | 06:25 | |
| Control Test | | 0.080 | 06:26 | |
| Air Blank | | 0.000 | 06:26 | |
| Diagnostics Check | | OK | 06:26 | |

Data Download Date/Time: 08/28/2019 19:29

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001293

INSTRUMENT REGISTERED TO: WILLISTON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,MARQUIS,A
BREATH TEST OPERATOR AGENCY: Williston PD
ARRESTING OFFICER: WRIGHT,MARQUIS
ARRESTING OFFICER AGENCY: Williston PD
LAST AGENCY INSPECTION: 08/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 937240
DRY GAS EXPIRY DATE: 04/07/2021
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: MATOS

DATE/TIME OF TEST: 08/31/2019 00:30:43

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:24 |
| Air Blank | | 0.000 | 00:25 |
| Control Test | | 0.078 | 00:25 |
| Air Blank | | 0.000 | 00:25 |
| Subject Sample | # 1 | 0.222 | 00:26 |
| Breath Volume | | 2.117 | 00:26 |
| Air Blank | | 0.000 | 00:26 |
| Air Blank | | 0.000 | 00:28 |
| Subject Sample | # 2 | 0.224 | 00:29 |
| Breath Volume | | 1.664 | 00:29 |
| Air Blank | | 0.000 | 00:29 |
| Control Test | | 0.078 | 00:30 |
| Air Blank | | 0.000 | 00:30 |
| Diagnostics Check | | OK | 00:30 |

Data Download Date/Time: 09/26/2019 16:42

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARVOTH,DANIEL,P
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: HARVOTH,DANIEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1141545
DRY GAS EXPIRY DATE: 05/11/2022
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: MOSS

DATE/TIME OF TEST: 08/24/2019 02:09:09

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 02:00 | |
| Air Blank | | 0.000 | 02:01 | |
| Control Test | | 0.079 | 02:01 | |
| Air Blank | | 0.000 | 02:01 | |
| Subject Sample | # 1 | | 02:05 | No Sample Provided |
| Breath Volume | | 0.000 | 02:05 | |
| Air Blank | | 0.000 | 02:05 | |
| Air Blank | | 0.000 | 02:07 | |
| Subject Sample | # 2 | | 02:07 | Subject Test Refused |
| Breath Volume | | 0.000 | 02:07 | |
| Air Blank | | 0.000 | 02:08 | |
| Control Test | | 0.080 | 02:08 | |
| Air Blank | | 0.000 | 02:09 | |
| Diagnostics Check | | OK | 02:09 | |

Data Download Date/Time: 09/24/2019 20:09

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOLLIVER,JORDON,P
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: TOLLIVER,JORDON
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1141545
DRY GAS EXPIRY DATE: 05/11/2022
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: MCCALVIN

DATE/TIME OF TEST: 08/26/2019 01:33:41

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:27 |
| Air Blank | | 0.000 | 01:27 |
| Control Test | | 0.080 | 01:27 |
| Air Blank | | 0.000 | 01:28 |
| Subject Sample | # 1 | 0.000 | 01:29 |
| Breath Volume | | 1.496 | 01:29 |
| Air Blank | | 0.000 | 01:29 |
| Air Blank | | 0.000 | 01:31 |
| Subject Sample | # 2 | 0.000 | 01:32 |
| Breath Volume | | 1.417 | 01:32 |
| Air Blank | | 0.000 | 01:32 |
| Control Test | | 0.080 | 01:33 |
| Air Blank | | 0.000 | 01:33 |
| Diagnostics Check | | OK | 01:33 |

Data Download Date/Time: 09/24/2019 20:09

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARVOTH,DANIEL,P
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: HARVOTH,DANIEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1141545
DRY GAS EXPIRY DATE: 05/11/2022
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 08/28/2019 23:54:54

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 23:51 | |
| Air Blank | | 0.000 | 23:51 | |
| Control Test | | 0.080 | 23:51 | |
| Air Blank | | 0.000 | 23:52 | |
| Subject Sample | # 1 | | 23:53 | Subject Test Refused |
| Breath Volume | | 0.000 | 23:53 | |
| Air Blank | | 0.000 | 23:54 | |
| Control Test | | 0.081 | 23:54 | |
| Air Blank | | 0.000 | 23:54 | |
| Diagnostics Check | | OK | 23:54 | |

Data Download Date/Time: 09/24/2019 20:09

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER,RAYMOND,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SPENCER,RAYMOND
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1141545
DRY GAS EXPIRY DATE: 05/11/2022
VIOLATION CODE: DUI
ARREST TIME: 22:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:01

SUBJECT'S LAST NAME: PENNINGTON

DATE/TIME OF TEST: 08/31/2019 23:31:48

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:21 |
| Air Blank | | 0.000 | 23:22 |
| Control Test | | 0.079 | 23:22 |
| Air Blank | | 0.000 | 23:23 |
| Subject Sample | # 1 | 0.273 | 23:24 |
| Breath Volume | | 2.953 | 23:24 |
| Air Blank | | 0.000 | 23:24 |
| Air Blank | | 0.000 | 23:26 |
| Subject Sample | # 2 | 0.243 | 23:27 |
| Breath Volume | | 1.171 | 23:27 |
| Air Blank | | 0.000 | 23:27 |
| Air Blank | | 0.000 | 23:29 |
| Subject Sample | # 3 | 0.277 | 23:30 |
| Breath Volume | | 3.312 | 23:30 |
| Air Blank | | 0.000 | 23:30 |
| Control Test | | 0.079 | 23:31 |
| Air Blank | | 0.000 | 23:31 |
| Diagnostics Check | | OK | 23:31 |

Data Download Date/Time: 09/24/2019 20:09

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER,RAYMOND,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SPENCER,RAYMOND
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1141545
DRY GAS EXPIRY DATE: 05/11/2022
VIOLATION CODE: DUI
ARREST TIME: 22:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:01

SUBJECT'S LAST NAME: PENNINGTON

DATE/TIME OF TEST: 08/31/2019 23:40:35

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:33 |
| Air Blank | | 0.000 | 23:34 |
| Control Test | | 0.078 | 23:34 |
| Air Blank | | 0.000 | 23:35 |
| Subject Sample | # 1 | 0.267 | 23:35 |
| Breath Volume | | 3.277 | 23:35 |
| Air Blank | | 0.000 | 23:36 |
| Air Blank | | 0.000 | 23:38 |
| Subject Sample | # 2 | 0.269 | 23:39 |
| Breath Volume | | 2.902 | 23:39 |
| Air Blank | | 0.000 | 23:39 |
| Control Test | | 0.079 | 23:39 |
| Air Blank | | 0.000 | 23:40 |
| Diagnostics Check | | OK | 23:40 |

Data Download Date/Time: 09/24/2019 20:09

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFITH,DALTON,M
BREATH TEST OPERATOR AGENCY: Marianna PD
ARRESTING OFFICER: GRAMLING,BRANDON
ARRESTING OFFICER AGENCY: Marianna PD
LAST AGENCY INSPECTION: 07/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 08/03/2019 22:42:21

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 22:35 |
| Air Blank | | 0.000 | 22:36 |
| Control Test | | 0.082 | 22:36 |
| Air Blank | | 0.000 | 22:36 |
| Subject Sample | # 1 | 0.131 | 22:37 |
| Breath Volume | | 1.230 | 22:37 |
| Air Blank | | 0.000 | 22:38 |
| Air Blank | | 0.000 | 22:39 |
| Subject Sample | # 2 | 0.129 | 22:40 |
| Breath Volume | | 1.906 | 22:40 |
| Air Blank | | 0.000 | 22:41 |
| Control Test | | 0.082 | 22:41 |
| Air Blank | | 0.000 | 22:42 |
| Diagnostics Check | | OK | 22:42 |

Data Download Date/Time: 08/23/2019 16:19

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001307

INSTRUMENT REGISTERED TO: MADISON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,JOSHUA,N
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: WRIGHT,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 08/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A3
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 12:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:24

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 08/24/2019 14:52:56

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 14:47 | |
| Air Blank | | 0.000 | 14:47 | |
| Control Test | | 0.075 | 14:47 | |
| Air Blank | | 0.000 | 14:48 | |
| Subject Sample | # 1 | 0.000 | 14:48 | |
| Breath Volume | | 3.601 | 14:48 | |
| Air Blank | | 0.000 | 14:49 | |
| Air Blank | | 0.000 | 14:51 | |
| Subject Sample | # 2 | 0.000 | 14:51 | |
| Breath Volume | | 3.234 | 14:51 | |
| Air Blank | | 0.000 | 14:52 | |
| Control Test | | 0.074 | 14:52 | Control Outside Tolerance |
| Air Blank | | 0.000 | 14:52 | |

Data Download Date/Time: 09/05/2019 11:26

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MERRITT,DAKOTA,A
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: MERRITT,DAKOTA
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 07/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: WAHL

DATE/TIME OF TEST: 08/03/2019 22:38:04

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 22:29 |
| Air Blank | | 0.000 | 22:30 |
| Control Test | | 0.081 | 22:30 |
| Air Blank | | 0.000 | 22:31 |
| Subject Sample | # 1 | 0.198 | 22:33 |
| Breath Volume | | 2.011 | 22:33 |
| Air Blank | | 0.000 | 22:33 |
| Air Blank | | 0.000 | 22:35 |
| Subject Sample | # 2 | 0.191 | 22:36 |
| Breath Volume | | 1.171 | 22:36 |
| Air Blank | | 0.000 | 22:37 |
| Control Test | | 0.079 | 22:37 |
| Air Blank | | 0.000 | 22:37 |
| Diagnostics Check | | OK | 22:38 |

Data Download Date/Time: 08/30/2019 14:22

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BASFORD,NICOLE,M
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: BASFORD,NICOLE
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 07/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 13:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:43

SUBJECT'S LAST NAME: SHARP

DATE/TIME OF TEST: 08/05/2019 14:39:44

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 14:33 |
| Air Blank | | 0.000 | 14:34 |
| Control Test | | 0.080 | 14:34 |
| Air Blank | | 0.000 | 14:35 |
| Subject Sample | # 1 | 0.000 | 14:35 |
| Breath Volume | | 1.304 | 14:35 |
| Air Blank | | 0.000 | 14:36 |
| Air Blank | | 0.000 | 14:38 |
| Subject Sample | # 2 | 0.000 | 14:38 |
| Breath Volume | | 1.292 | 14:38 |
| Air Blank | | 0.000 | 14:38 |
| Control Test | | 0.080 | 14:39 |
| Air Blank | | 0.000 | 14:39 |
| Diagnostics Check | | OK | 14:39 |

Data Download Date/Time: 08/30/2019 14:22

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,SARAH,E
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: SMITH,SARAH
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 07/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: ADMIN
ARREST TIME: 15:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:45

SUBJECT'S LAST NAME: BRAMBLETTE

DATE/TIME OF TEST: 08/16/2019 16:11:21

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 16:05 |
| Air Blank | | 0.000 | 16:05 |
| Control Test | | 0.081 | 16:06 |
| Air Blank | | 0.000 | 16:06 |
| Subject Sample | # 1 | 0.016 | 16:07 |
| Breath Volume | | 1.472 | 16:07 |
| Air Blank | | 0.000 | 16:07 |
| Air Blank | | 0.000 | 16:09 |
| Subject Sample | # 2 | 0.014 | 16:09 |
| Breath Volume | | 1.402 | 16:09 |
| Air Blank | | 0.000 | 16:10 |
| Control Test | | 0.080 | 16:10 |
| Air Blank | | 0.000 | 16:11 |
| Diagnostics Check | | OK | 16:11 |

Data Download Date/Time: 08/30/2019 14:22

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,SARAH,E
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: CHASTAIN,KRISTA
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 07/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 13:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:29

SUBJECT'S LAST NAME: PADGETT

DATE/TIME OF TEST: 08/26/2019 16:31:37

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 16:23 | |
| Air Blank | | 0.000 | 16:24 | |
| Control Test | | 0.081 | 16:24 | |
| Air Blank | | 0.000 | 16:25 | |
| Subject Sample | # 1 | | 16:28 | No Sample Provided |
| Breath Volume | | 0.000 | 16:28 | |
| Air Blank | | 0.000 | 16:28 | |
| Air Blank | | 0.000 | 16:30 | |
| Subject Sample | # 2 | | 16:31 | Sequence Aborted |
| Breath Volume | | 0.000 | 16:31 | |
| Air Blank | | 0.000 | 16:31 | |

Data Download Date/Time: 08/30/2019 14:22

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,SARAH,E
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: SMITH,SARAH
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 08/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: ADMIN
ARREST TIME: 15:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:57

SUBJECT'S LAST NAME: BRAMBLETTE

DATE/TIME OF TEST: 08/30/2019 16:23:35

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 16:17 |
| Air Blank | | 0.000 | 16:18 |
| Control Test | | 0.081 | 16:18 |
| Air Blank | | 0.000 | 16:19 |
| Subject Sample | # 1 | 0.000 | 16:19 |
| Breath Volume | | 1.488 | 16:19 |
| Air Blank | | 0.000 | 16:19 |
| Air Blank | | 0.000 | 16:21 |
| Subject Sample | # 2 | 0.000 | 16:22 |
| Breath Volume | | 1.535 | 16:22 |
| Air Blank | | 0.000 | 16:22 |
| Control Test | | 0.081 | 16:23 |
| Air Blank | | 0.000 | 16:23 |
| Diagnostics Check | | OK | 16:23 |

Data Download Date/Time: 09/30/2019 19:30

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,SARAH,E
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: SMITH,SARAH
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 08/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: SYS CHECK
ARREST TIME: 16:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:58

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 08/30/2019 17:25:03

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 17:19 |
| Air Blank | | 0.000 | 17:19 |
| Control Test | | 0.080 | 17:19 |
| Air Blank | | 0.000 | 17:20 |
| Subject Sample | # 1 | 0.000 | 17:20 |
| Breath Volume | | 1.710 | 17:20 |
| Air Blank | | 0.000 | 17:21 |
| Air Blank | | 0.000 | 17:23 |
| Subject Sample | # 2 | 0.000 | 17:23 |
| Breath Volume | | 1.300 | 17:23 |
| Air Blank | | 0.000 | 17:24 |
| Control Test | | 0.082 | 17:24 |
| Air Blank | | 0.000 | 17:24 |
| Diagnostics Check | | OK | 17:25 |

Data Download Date/Time: 09/30/2019 19:30

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOBEL,EARL,L
BREATH TEST OPERATOR AGENCY: Mexico Beach PD
ARRESTING OFFICER: KNOBEL,EARL
ARRESTING OFFICER AGENCY: Mexico Beach PD
LAST AGENCY INSPECTION: 07/31/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: STANSEL

DATE/TIME OF TEST: 08/16/2019 23:20:08

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 23:15 | |
| Air Blank | | 0.000 | 23:15 | |
| Control Test | | 0.079 | 23:16 | |
| Air Blank | | 0.000 | 23:16 | |
| Subject Sample | # 1 | | 23:18 | Subject Test Refused |
| Breath Volume | | 0.000 | 23:18 | |
| Air Blank | | 0.000 | 23:19 | |
| Control Test | | 0.079 | 23:19 | |
| Air Blank | | 0.000 | 23:20 | |
| Diagnostics Check | | OK | 23:20 | |

Data Download Date/Time: 08/30/2019 14:14

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOBEL,EARL,L
BREATH TEST OPERATOR AGENCY: Mexico Beach PD
ARRESTING OFFICER: KNOBEL,EARL
ARRESTING OFFICER AGENCY: Mexico Beach PD
LAST AGENCY INSPECTION: 07/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 04:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:35

SUBJECT'S LAST NAME: ELLIOTT

DATE/TIME OF TEST: 08/17/2019 06:10:34

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 05:58 | |
| Air Blank | | 0.000 | 05:59 | |
| Control Test | | 0.080 | 05:59 | |
| Air Blank | | 0.000 | 05:59 | |
| Subject Sample | # 1 | 0.085 | 06:03 | Volume Not Met |
| Breath Volume | | 0.132 | 06:03 | |
| Air Blank | | 0.000 | 06:03 | |
| Air Blank | | 0.000 | 06:05 | |
| Subject Sample | # 2 | 0.082 | 06:06 | |
| Breath Volume | | 1.410 | 06:06 | |
| Air Blank | | 0.000 | 06:06 | |
| Air Blank | | 0.000 | 06:08 | |
| Subject Sample | # 3 | 0.079 | 06:09 | |
| Breath Volume | | 2.105 | 06:09 | |
| Air Blank | | 0.000 | 06:09 | |
| Control Test | | 0.078 | 06:09 | |
| Air Blank | | 0.000 | 06:10 | |
| Diagnostics Check | | OK | 06:10 | |

Data Download Date/Time: 08/30/2019 14:14

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEWBY,JOHN,O
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: RICHARD,MICHAEL
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 07/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 869968
DRY GAS EXPIRY DATE: 12/06/2020
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: TABB

DATE/TIME OF TEST: 08/10/2019 05:03:43

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 05:00 | |
| Air Blank | | 0.000 | 05:01 | |
| Control Test | | 0.079 | 05:01 | |
| Air Blank | | 0.000 | 05:02 | |
| Subject Sample | # 1 | | 05:02 | Subject Test Refused |
| Breath Volume | | 0.000 | 05:02 | |
| Air Blank | | 0.000 | 05:02 | |
| Control Test | | 0.080 | 05:03 | |
| Air Blank | | 0.000 | 05:03 | |
| Diagnostics Check | | OK | 05:03 | |

Data Download Date/Time: 08/15/2019 11:24

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POVEDA,CARLOS,M
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: FORSMAN,MICHAEL
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 07/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 869968
DRY GAS EXPIRY DATE: 12/06/2020
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: LACY

DATE/TIME OF TEST: 08/13/2019 01:10:37

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:04 |
| Air Blank | | 0.000 | 01:04 |
| Control Test | | 0.079 | 01:05 |
| Air Blank | | 0.000 | 01:05 |
| Subject Sample | # 1 | 0.198 | 01:06 |
| Breath Volume | | 1.761 | 01:06 |
| Air Blank | | 0.000 | 01:06 |
| Air Blank | | 0.000 | 01:08 |
| Subject Sample | # 2 | 0.197 | 01:09 |
| Breath Volume | | 1.753 | 01:09 |
| Air Blank | | 0.000 | 01:09 |
| Control Test | | 0.079 | 01:10 |
| Air Blank | | 0.000 | 01:10 |
| Diagnostics Check | | OK | 01:10 |

Data Download Date/Time: 08/15/2019 11:24

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DANIEL,R
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: HALL,DANIEL
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 08/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 869968
DRY GAS EXPIRY DATE: 12/06/2020
VIOLATION CODE: DUI
ARREST TIME: 09:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:28

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 08/25/2019 10:10:11

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 10:03 |
| Air Blank | | 0.000 | 10:03 |
| Control Test | | 0.079 | 10:03 |
| Air Blank | | 0.000 | 10:04 |
| Subject Sample | # 1 | 0.000 | 10:05 |
| Breath Volume | | 1.718 | 10:05 |
| Air Blank | | 0.000 | 10:05 |
| Air Blank | | 0.000 | 10:07 |
| Subject Sample | # 2 | 0.000 | 10:08 |
| Breath Volume | | 2.105 | 10:08 |
| Air Blank | | 0.000 | 10:09 |
| Control Test | | 0.079 | 10:09 |
| Air Blank | | 0.000 | 10:10 |
| Diagnostics Check | | OK | 10:10 |

Data Download Date/Time: 09/11/2019 18:02

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRAN,DAVID,V
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: SAENZ,MELLISSA
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 08/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 869968
DRY GAS EXPIRY DATE: 12/06/2020
VIOLATION CODE: DUI
ARREST TIME: 07:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:34

SUBJECT'S LAST NAME: STONE

DATE/TIME OF TEST: 08/27/2019 09:09:20

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 09:03 |
| Air Blank | | 0.000 | 09:03 |
| Control Test | | 0.079 | 09:04 |
| Air Blank | | 0.000 | 09:04 |
| Subject Sample | # 1 | 0.000 | 09:05 |
| Breath Volume | | 3.554 | 09:05 |
| Air Blank | | 0.000 | 09:05 |
| Air Blank | | 0.000 | 09:07 |
| Subject Sample | # 2 | 0.000 | 09:07 |
| Breath Volume | | 1.691 | 09:07 |
| Air Blank | | 0.000 | 09:08 |
| Control Test | | 0.080 | 09:08 |
| Air Blank | | 0.000 | 09:09 |
| Diagnostics Check | | OK | 09:09 |

Data Download Date/Time: 09/11/2019 18:02

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNGHANSE,KYLE,O
BREATH TEST OPERATOR AGENCY: DeFuniak Springs PD
ARRESTING OFFICER: YOUNGHANSE,KYLE
ARRESTING OFFICER AGENCY: DeFuniak Springs PD
LAST AGENCY INSPECTION: 07/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: SELLARS

DATE/TIME OF TEST: 08/01/2019 00:20:31

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 00:07 | |
| Air Blank | | 0.000 | 00:07 | |
| Control Test | | 0.080 | 00:08 | |
| Air Blank | | 0.000 | 00:08 | |
| Subject Sample | # 1 | 0.352 | 00:11 | Volume Not Met |
| Breath Volume | | 0.464 | 00:11 | |
| Air Blank | | 0.000 | 00:12 | |
| Air Blank | | 0.000 | 00:14 | |
| Subject Sample | # 2 | 0.341 | 00:15 | No .020 agreement |
| Breath Volume | | 1.742 | 00:15 | |
| Air Blank | | 0.000 | 00:16 | |
| Air Blank | | 0.000 | 00:17 | |
| Subject Sample | # 3 | 0.317 | 00:18 | No .020 agreement |
| Breath Volume | | 1.523 | 00:18 | |
| Air Blank | | 0.000 | 00:19 | |
| Control Test | | 0.078 | 00:19 | |
| Air Blank | | 0.000 | 00:20 | |
| Diagnostics Check | | OK | 00:20 | |

Data Download Date/Time: 08/25/2019 22:14

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLISON,STEVEN,L
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: LYTLE,ALLI
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 07/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: LUKE

DATE/TIME OF TEST: 08/01/2019 04:45:55

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 04:36 | |
| Air Blank | | 0.000 | 04:36 | |
| Control Test | | 0.079 | 04:36 | |
| Air Blank | | 0.000 | 04:37 | |
| Subject Sample | # 1 | 0.182 | 04:40 | Volume Not Met |
| Breath Volume | | 0.144 | 04:40 | |
| Air Blank | | 0.000 | 04:41 | |
| Air Blank | | 0.000 | 04:42 | |
| Subject Sample | # 2 | | 04:44 | Subject Test Refused |
| Breath Volume | | 0.281 | 04:44 | |
| Air Blank | | 0.000 | 04:44 | |
| Control Test | | 0.077 | 04:45 | |
| Air Blank | | 0.000 | 04:45 | |
| Diagnostics Check | | OK | 04:45 | |

Data Download Date/Time: 08/25/2019 22:14

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENNING,LUKE,
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: BUDUEN,PABLO
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 07/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 15:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:20

SUBJECT'S LAST NAME: VAN EATON

DATE/TIME OF TEST: 08/07/2019 16:49:04

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 16:45 | |
| Air Blank | | 0.000 | 16:45 | |
| Control Test | | 0.080 | 16:46 | |
| Air Blank | | 0.000 | 16:46 | |
| Subject Sample | # 1 | | 16:47 | Subject Test Refused |
| Breath Volume | | 0.000 | 16:47 | |
| Air Blank | | 0.000 | 16:48 | |
| Control Test | | 0.080 | 16:48 | |
| Air Blank | | 0.000 | 16:48 | |
| Diagnostics Check | | OK | 16:49 | |

Data Download Date/Time: 08/25/2019 22:14

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAHAN,WILLIAM,D
BREATH TEST OPERATOR AGENCY: DeFuniak Springs PD
ARRESTING OFFICER: PAUL,WILLIAM
ARRESTING OFFICER AGENCY: DeFuniak Springs PD
LAST AGENCY INSPECTION: 07/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:08

SUBJECT'S LAST NAME: SASSER

DATE/TIME OF TEST: 08/08/2019 21:34:51

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 21:28 |
| Air Blank | | 0.000 | 21:29 |
| Control Test | | 0.080 | 21:29 |
| Air Blank | | 0.000 | 21:30 |
| Subject Sample | # 1 | 0.084 | 21:30 |
| Breath Volume | | 1.558 | 21:30 |
| Air Blank | | 0.000 | 21:31 |
| Air Blank | | 0.000 | 21:32 |
| Subject Sample | # 2 | 0.080 | 21:33 |
| Breath Volume | | 1.218 | 21:33 |
| Air Blank | | 0.000 | 21:33 |
| Control Test | | 0.077 | 21:34 |
| Air Blank | | 0.000 | 21:34 |
| Diagnostics Check | | OK | 21:34 |

Data Download Date/Time: 08/25/2019 22:15

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FUGATE,MITCHELL,W
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: YOST,CRAIG
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 07/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 19:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:34

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 08/11/2019 21:14:28

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 21:03 | |
| Air Blank | | 0.000 | 21:03 | |
| Control Test | | 0.080 | 21:03 | |
| Air Blank | | 0.000 | 21:04 | |
| Subject Sample | # 1 | | 21:07 | No Sample Provided |
| Breath Volume | | 0.000 | 21:07 | |
| Air Blank | | 0.000 | 21:07 | |
| Air Blank | | 0.000 | 21:09 | |
| Subject Sample | # 2 | | 21:13 | No Sample Provided |
| Breath Volume | | 0.000 | 21:13 | |
| Air Blank | | 0.000 | 21:13 | |
| Control Test | | 0.080 | 21:13 | |
| Air Blank | | 0.000 | 21:14 | |
| Diagnostics Check | | OK | 21:14 | |

Data Download Date/Time: 08/25/2019 22:15

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNGHANSE,KYLE,O
BREATH TEST OPERATOR AGENCY: DeFuniak Springs PD
ARRESTING OFFICER: YOUNGHANSE,KYLE
ARRESTING OFFICER AGENCY: DeFuniak Springs PD
LAST AGENCY INSPECTION: 07/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 18:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: FARROW

DATE/TIME OF TEST: 08/15/2019 19:39:54

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 19:37 | |
| Air Blank | | 0.000 | 19:37 | |
| Control Test | | 0.080 | 19:37 | |
| Air Blank | | 0.000 | 19:38 | |
| Subject Sample | # 1 | | 19:38 | Subject Test Refused |
| Breath Volume | | 0.000 | 19:38 | |
| Air Blank | | 0.000 | 19:38 | |
| Control Test | | 0.080 | 19:39 | |
| Air Blank | | 0.000 | 19:39 | |
| Diagnostics Check | | OK | 19:39 | |

Data Download Date/Time: 08/25/2019 22:15

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNGHANSE,KYLE,O
BREATH TEST OPERATOR AGENCY: DeFuniak Springs PD
ARRESTING OFFICER: LEWIS,PAUL
ARRESTING OFFICER AGENCY: DeFuniak Springs PD
LAST AGENCY INSPECTION: 07/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: DUPONT

DATE/TIME OF TEST: 08/16/2019 00:20:47

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 00:17 | |
| Air Blank | | 0.000 | 00:18 | |
| Control Test | | 0.080 | 00:18 | |
| Air Blank | | 0.000 | 00:19 | |
| Subject Sample | # 1 | | 00:19 | Subject Test Refused |
| Breath Volume | | 0.000 | 00:19 | |
| Air Blank | | 0.000 | 00:19 | |
| Control Test | | 0.079 | 00:20 | |
| Air Blank | | 0.000 | 00:20 | |
| Diagnostics Check | | OK | 00:20 | |

Data Download Date/Time: 08/25/2019 22:15

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVERY,JOHN,W
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: PEOPLES,NATHAN
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 07/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 19:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: COWDEN

DATE/TIME OF TEST: 08/21/2019 20:39:34

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 20:33 |
| Air Blank | | 0.000 | 20:33 |
| Control Test | | 0.079 | 20:33 |
| Air Blank | | 0.000 | 20:34 |
| Subject Sample | # 1 | 0.393 | 20:34 |
| Breath Volume | | 2.480 | 20:34 |
| Air Blank | | 0.000 | 20:35 |
| Air Blank | | 0.000 | 20:37 |
| Subject Sample | # 2 | 0.403 | 20:37 |
| Breath Volume | | 2.281 | 20:37 |
| Air Blank | | 0.000 | 20:38 |
| Control Test | | 0.078 | 20:38 |
| Air Blank | | 0.000 | 20:39 |
| Diagnostics Check | | OK | 20:39 |

Data Download Date/Time: 08/25/2019 22:15

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVERY,JOHN,W
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: PEOPLES,NATHAN
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 07/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 08/23/2019 03:56:47

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 03:53 | |
| Air Blank | | 0.000 | 03:54 | |
| Control Test | | 0.079 | 03:54 | |
| Air Blank | | 0.000 | 03:55 | |
| Subject Sample | # 1 | | 03:55 | Subject Test Refused |
| Breath Volume | | 0.000 | 03:55 | |
| Air Blank | | 0.000 | 03:55 | |
| Control Test | | 0.080 | 03:56 | |
| Air Blank | | 0.000 | 03:56 | |
| Diagnostics Check | | OK | 03:56 | |

Data Download Date/Time: 08/25/2019 22:15

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVERY,JOHN,W
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: BRANNON,JUSTIN
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: COFFEE

DATE/TIME OF TEST: 08/30/2019 22:22:18

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 22:19 | |
| Air Blank | | 0.000 | 22:19 | |
| Control Test | | 0.079 | 22:20 | |
| Air Blank | | 0.000 | 22:20 | |
| Subject Sample | # 1 | | 22:20 | Subject Test Refused |
| Breath Volume | | 0.000 | 22:20 | |
| Air Blank | | 0.000 | 22:21 | |
| Control Test | | 0.079 | 22:21 | |
| Air Blank | | 0.000 | 22:22 | |
| Diagnostics Check | | OK | 22:22 | |

Data Download Date/Time: 09/26/2019 11:53

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHULTZ,JEREMY,J
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: WHITAKER,JOHNNY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: DEAN

DATE/TIME OF TEST: 08/04/2019 22:48:53

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 22:42 |
| Air Blank | | 0.000 | 22:42 |
| Control Test | | 0.080 | 22:43 |
| Air Blank | | 0.000 | 22:43 |
| Subject Sample | # 1 | 0.182 | 22:44 |
| Breath Volume | | 3.781 | 22:44 |
| Air Blank | | 0.000 | 22:45 |
| Air Blank | | 0.000 | 22:46 |
| Subject Sample | # 2 | 0.182 | 22:47 |
| Breath Volume | | 3.128 | 22:47 |
| Air Blank | | 0.000 | 22:47 |
| Control Test | | 0.078 | 22:48 |
| Air Blank | | 0.000 | 22:48 |
| Diagnostics Check | | OK | 22:48 |

Data Download Date/Time: 08/25/2019 22:28

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCORMICK,CHANTEL,B
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: MCCORMICK,CHANTEL
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 07/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 14:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:40

SUBJECT'S LAST NAME: POWELL

DATE/TIME OF TEST: 08/06/2019 16:05:13

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 16:02 | |
| Air Blank | | 0.000 | 16:02 | |
| Control Test | | 0.079 | 16:03 | |
| Air Blank | | 0.000 | 16:03 | |
| Subject Sample | # 1 | | 16:03 | Subject Test Refused |
| Breath Volume | | 0.000 | 16:03 | |
| Air Blank | | 0.000 | 16:04 | |
| Control Test | | 0.080 | 16:04 | |
| Air Blank | | 0.000 | 16:05 | |
| Diagnostics Check | | OK | 16:05 | |

Data Download Date/Time: 08/25/2019 22:28

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PULLINGS,ALLEN,C
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: PULLINGS,ALLEN
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 07/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: PROB/PAROL
ARREST TIME: 22:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 08/14/2019 23:54:44

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:48 |
| Air Blank | | 0.000 | 23:48 |
| Control Test | | 0.080 | 23:49 |
| Air Blank | | 0.000 | 23:49 |
| Subject Sample | # 1 | 0.170 | 23:50 |
| Breath Volume | | 1.429 | 23:50 |
| Air Blank | | 0.000 | 23:50 |
| Air Blank | | 0.000 | 23:52 |
| Subject Sample | # 2 | 0.168 | 23:53 |
| Breath Volume | | 1.757 | 23:53 |
| Air Blank | | 0.000 | 23:53 |
| Control Test | | 0.081 | 23:54 |
| Air Blank | | 0.000 | 23:54 |
| Diagnostics Check | | OK | 23:54 |

Data Download Date/Time: 08/25/2019 22:28

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUNN,CONNOR,S
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: TICE,HUNTER
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 07/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 10:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:18

SUBJECT'S LAST NAME: BEGAULT

DATE/TIME OF TEST: 08/16/2019 10:59:09

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 10:52 |
| Air Blank | | 0.000 | 10:53 |
| Control Test | | 0.079 | 10:53 |
| Air Blank | | 0.000 | 10:54 |
| Subject Sample | # 1 | 0.000 | 10:54 |
| Breath Volume | | 2.464 | 10:54 |
| Air Blank | | 0.000 | 10:55 |
| Air Blank | | 0.000 | 10:57 |
| Subject Sample | # 2 | 0.000 | 10:57 |
| Breath Volume | | 2.261 | 10:57 |
| Air Blank | | 0.000 | 10:58 |
| Control Test | | 0.079 | 10:58 |
| Air Blank | | 0.000 | 10:59 |
| Diagnostics Check | | OK | 10:59 |

Data Download Date/Time: 08/25/2019 22:28

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHULTZ,JEREMY,J
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: SCHULTZ,JEREMY
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 07/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: POLZIN

DATE/TIME OF TEST: 08/23/2019 01:30:50

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:24 |
| Air Blank | | 0.000 | 01:24 |
| Control Test | | 0.080 | 01:24 |
| Air Blank | | 0.000 | 01:25 |
| Subject Sample | # 1 | 0.189 | 01:26 |
| Breath Volume | | 3.441 | 01:26 |
| Air Blank | | 0.000 | 01:26 |
| Air Blank | | 0.000 | 01:28 |
| Subject Sample | # 2 | 0.180 | 01:29 |
| Breath Volume | | 3.093 | 01:29 |
| Air Blank | | 0.000 | 01:29 |
| Control Test | | 0.078 | 01:30 |
| Air Blank | | 0.000 | 01:30 |
| Diagnostics Check | | OK | 01:30 |

Data Download Date/Time: 08/25/2019 22:29

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PULLINGS,ALLEN,C
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: HATLEY,KARAGAN
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 07/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: PORTER

DATE/TIME OF TEST: 08/24/2019 23:27:57

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 23:16 | |
| Air Blank | | 0.000 | 23:17 | |
| Control Test | | 0.079 | 23:17 | |
| Air Blank | | 0.000 | 23:17 | |
| Subject Sample | # 1 | | 23:20 | No Sample Provided |
| Breath Volume | | 0.000 | 23:20 | |
| Air Blank | | 0.000 | 23:21 | |
| Air Blank | | 0.000 | 23:23 | |
| Subject Sample | # 2 | | 23:26 | No Sample Provided |
| Breath Volume | | 0.000 | 23:26 | |
| Air Blank | | 0.000 | 23:27 | |
| Control Test | | 0.081 | 23:27 | |
| Air Blank | | 0.000 | 23:27 | |
| Diagnostics Check | | OK | 23:27 | |

Data Download Date/Time: 08/25/2019 22:29

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHULTZ,JEREMY,J
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: SCHULTZ,JEREMY
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: MALL

DATE/TIME OF TEST: 08/26/2019 21:45:54

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 21:39 |
| Air Blank | | 0.000 | 21:39 |
| Control Test | | 0.080 | 21:40 |
| Air Blank | | 0.000 | 21:40 |
| Subject Sample | # 1 | 0.009 | 21:41 |
| Breath Volume | | 2.949 | 21:41 |
| Air Blank | | 0.000 | 21:42 |
| Air Blank | | 0.000 | 21:44 |
| Subject Sample | # 2 | 0.008 | 21:44 |
| Breath Volume | | 3.128 | 21:44 |
| Air Blank | | 0.000 | 21:44 |
| Control Test | | 0.078 | 21:45 |
| Air Blank | | 0.000 | 21:45 |
| Diagnostics Check | | OK | 21:45 |

Data Download Date/Time: 09/26/2019 11:49

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHULTZ,JEREMY,J
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: VENTRY,WILLIS
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:06

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 08/30/2019 22:41:38

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 22:35 |
| Air Blank | | 0.000 | 22:35 |
| Control Test | | 0.080 | 22:36 |
| Air Blank | | 0.000 | 22:36 |
| Subject Sample | # 1 | 0.192 | 22:37 |
| Breath Volume | | 2.527 | 22:37 |
| Air Blank | | 0.000 | 22:37 |
| Air Blank | | 0.000 | 22:39 |
| Subject Sample | # 2 | 0.185 | 22:40 |
| Breath Volume | | 2.484 | 22:40 |
| Air Blank | | 0.000 | 22:40 |
| Control Test | | 0.079 | 22:41 |
| Air Blank | | 0.000 | 22:41 |
| Diagnostics Check | | OK | 22:41 |

Data Download Date/Time: 09/26/2019 11:49

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,MARK,A
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: VULETA,B
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 07/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 06:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:08

SUBJECT'S LAST NAME: CAN0-CASTILLO

DATE/TIME OF TEST: 08/03/2019 07:34:49

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 07:28 |
| Air Blank | | 0.000 | 07:29 |
| Control Test | | 0.081 | 07:29 |
| Air Blank | | 0.000 | 07:29 |
| Subject Sample | # 1 | 0.164 | 07:30 |
| Breath Volume | | 1.253 | 07:30 |
| Air Blank | | 0.000 | 07:31 |
| Air Blank | | 0.000 | 07:32 |
| Subject Sample | # 2 | 0.158 | 07:33 |
| Breath Volume | | 1.523 | 07:33 |
| Air Blank | | 0.000 | 07:33 |
| Control Test | | 0.081 | 07:34 |
| Air Blank | | 0.000 | 07:34 |
| Diagnostics Check | | OK | 07:34 |

Data Download Date/Time: 08/07/2019 09:01

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,MARK,A
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: VULETA,B
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 07/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 15:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:07

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 08/04/2019 16:33:21

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 16:27 |
| Air Blank | | 0.000 | 16:27 |
| Control Test | | 0.081 | 16:28 |
| Air Blank | | 0.000 | 16:28 |
| Subject Sample | # 1 | 0.000 | 16:29 |
| Breath Volume | | 3.117 | 16:29 |
| Air Blank | | 0.000 | 16:29 |
| Air Blank | | 0.000 | 16:31 |
| Subject Sample | # 2 | 0.000 | 16:31 |
| Breath Volume | | 3.328 | 16:31 |
| Air Blank | | 0.000 | 16:32 |
| Control Test | | 0.080 | 16:32 |
| Air Blank | | 0.000 | 16:33 |
| Diagnostics Check | | OK | 16:33 |

Data Download Date/Time: 08/07/2019 09:01

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODGERS,JAMES,E
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: LACERTE,JOEL
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 08/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: RODGERS

DATE/TIME OF TEST: 08/07/2019 22:00:17

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 21:56 | |
| Air Blank | | 0.000 | 21:56 | |
| Control Test | | 0.080 | 21:57 | |
| Air Blank | | 0.000 | 21:57 | |
| Subject Sample | # 1 | | 21:58 | Subject Test Refused |
| Breath Volume | | 0.000 | 21:58 | |
| Air Blank | | 0.000 | 21:59 | |
| Control Test | | 0.080 | 21:59 | |
| Air Blank | | 0.000 | 22:00 | |
| Diagnostics Check | | OK | 22:00 | |

Data Download Date/Time: 09/11/2019 10:12

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMMONS,RYANE,J
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: BOGNER,SHANE
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 08/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 12:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:20

SUBJECT'S LAST NAME: CYPRESS

DATE/TIME OF TEST: 08/09/2019 12:46:42

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 12:40 |
| Air Blank | | 0.000 | 12:40 |
| Control Test | | 0.079 | 12:41 |
| Air Blank | | 0.000 | 12:41 |
| Subject Sample | # 1 | 0.114 | 12:42 |
| Breath Volume | | 2.574 | 12:42 |
| Air Blank | | 0.000 | 12:42 |
| Air Blank | | 0.000 | 12:44 |
| Subject Sample | # 2 | 0.117 | 12:45 |
| Breath Volume | | 2.750 | 12:45 |
| Air Blank | | 0.000 | 12:45 |
| Control Test | | 0.080 | 12:46 |
| Air Blank | | 0.000 | 12:46 |
| Diagnostics Check | | OK | 12:46 |

Data Download Date/Time: 09/11/2019 10:12

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,STEVEN,D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK,STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 08/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: RAULERSON

DATE/TIME OF TEST: 08/10/2019 02:27:52

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 02:24 | |
| Air Blank | | 0.000 | 02:25 | |
| Control Test | | 0.080 | 02:25 | |
| Air Blank | | 0.000 | 02:26 | |
| Subject Sample | # 1 | | 02:26 | Subject Test Refused |
| Breath Volume | | 0.000 | 02:26 | |
| Air Blank | | 0.000 | 02:26 | |
| Control Test | | 0.080 | 02:27 | |
| Air Blank | | 0.000 | 02:27 | |
| Diagnostics Check | | OK | 02:27 | |

Data Download Date/Time: 09/11/2019 10:13

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,STEVEN,D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK,STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 08/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 18:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: BAIJNAUTH

DATE/TIME OF TEST: 08/10/2019 19:46:55

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 19:35 | |
| Air Blank | | 0.000 | 19:35 | |
| Control Test | | 0.081 | 19:36 | |
| Air Blank | | 0.000 | 19:36 | |
| Subject Sample | # 1 | 0.076 | 19:39 | Volume Not Met |
| Breath Volume | | 0.488 | 19:39 | |
| Air Blank | | 0.000 | 19:40 | |
| Air Blank | | 0.000 | 19:42 | |
| Subject Sample | # 2 | 0.077 | 19:45 | Volume Not Met |
| Breath Volume | | 0.421 | 19:45 | |
| Air Blank | | 0.000 | 19:45 | |
| Control Test | | 0.080 | 19:46 | |
| Air Blank | | 0.000 | 19:46 | |
| Diagnostics Check | | OK | 19:46 | |

Data Download Date/Time: 09/11/2019 10:13

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAUSLER,TREVOR,C
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: HAUSLER,TREVOR
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 08/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: BUI
ARREST TIME: 20:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: MELLETTE

DATE/TIME OF TEST: 08/10/2019 21:38:36

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 21:35 | |
| Air Blank | | 0.000 | 21:36 | |
| Control Test | | 0.080 | 21:36 | |
| Air Blank | | 0.000 | 21:36 | |
| Subject Sample | # 1 | | 21:37 | Subject Test Refused |
| Breath Volume | | 0.000 | 21:37 | |
| Air Blank | | 0.000 | 21:37 | |
| Control Test | | 0.080 | 21:38 | |
| Air Blank | | 0.000 | 21:38 | |
| Diagnostics Check | | OK | 21:38 | |

Data Download Date/Time: 09/11/2019 10:13

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAJERE,ARIC,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK,STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 08/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 08/11/2019 04:02:03

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:55 |
| Air Blank | | 0.000 | 03:56 |
| Control Test | | 0.080 | 03:56 |
| Air Blank | | 0.000 | 03:56 |
| Subject Sample | # 1 | 0.189 | 03:57 |
| Breath Volume | | 2.722 | 03:57 |
| Air Blank | | 0.000 | 03:58 |
| Air Blank | | 0.000 | 03:59 |
| Subject Sample | # 2 | 0.181 | 04:00 |
| Breath Volume | | 2.835 | 04:00 |
| Air Blank | | 0.000 | 04:01 |
| Control Test | | 0.079 | 04:01 |
| Air Blank | | 0.000 | 04:01 |
| Diagnostics Check | | OK | 04:02 |

Data Download Date/Time: 09/11/2019 10:13

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,STEVEN,D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK,STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 08/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: GALVAN

DATE/TIME OF TEST: 08/18/2019 00:33:07

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 00:30 | |
| Air Blank | | 0.000 | 00:30 | |
| Control Test | | 0.081 | 00:30 | |
| Air Blank | | 0.000 | 00:31 | |
| Subject Sample | # 1 | | 00:31 | Subject Test Refused |
| Breath Volume | | 0.000 | 00:31 | |
| Air Blank | | 0.000 | 00:32 | |
| Control Test | | 0.080 | 00:32 | |
| Air Blank | | 0.000 | 00:33 | |
| Diagnostics Check | | OK | 00:33 | |

Data Download Date/Time: 09/11/2019 10:13

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,DAVID,E
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK,STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 08/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: DEMPSEY

DATE/TIME OF TEST: 08/21/2019 22:07:44

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 22:00 |
| Air Blank | | 0.000 | 22:01 |
| Control Test | | 0.081 | 22:01 |
| Air Blank | | 0.000 | 22:02 |
| Subject Sample | # 1 | 0.139 | 22:03 |
| Breath Volume | | 3.046 | 22:03 |
| Air Blank | | 0.000 | 22:03 |
| Air Blank | | 0.000 | 22:05 |
| Subject Sample | # 2 | 0.142 | 22:06 |
| Breath Volume | | 3.210 | 22:06 |
| Air Blank | | 0.000 | 22:06 |
| Control Test | | 0.080 | 22:07 |
| Air Blank | | 0.000 | 22:07 |
| Diagnostics Check | | OK | 22:07 |

Data Download Date/Time: 09/11/2019 10:13

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,STEVEN,D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK,STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 08/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:03

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 08/23/2019 22:35:13

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 22:23 | |
| Air Blank | | 0.000 | 22:24 | |
| Control Test | | 0.080 | 22:24 | |
| Air Blank | | 0.000 | 22:24 | |
| Subject Sample | # 1 | | 22:28 | No Sample Provided |
| Breath Volume | | 0.000 | 22:28 | |
| Air Blank | | 0.000 | 22:28 | |
| Air Blank | | 0.000 | 22:30 | |
| Subject Sample | # 2 | | 22:33 | No Sample Provided |
| Breath Volume | | 0.000 | 22:33 | |
| Air Blank | | 0.000 | 22:34 | |
| Control Test | | 0.078 | 22:34 | |
| Air Blank | | 0.000 | 22:35 | |
| Diagnostics Check | | OK | 22:35 | |

Data Download Date/Time: 09/11/2019 10:13

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NICHOLSON,HORACE,O
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: LACERTE,JOEL
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 08/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 14:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:33

SUBJECT'S LAST NAME: RUTHERFORD

DATE/TIME OF TEST: 08/30/2019 15:00:20

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 14:53 |
| Air Blank | | 0.000 | 14:54 |
| Control Test | | 0.081 | 14:54 |
| Air Blank | | 0.000 | 14:55 |
| Subject Sample | # 1 | 0.000 | 14:55 |
| Breath Volume | | 2.445 | 14:55 |
| Air Blank | | 0.000 | 14:56 |
| Air Blank | | 0.000 | 14:58 |
| Subject Sample | # 2 | 0.000 | 14:58 |
| Breath Volume | | 2.390 | 14:58 |
| Air Blank | | 0.000 | 14:59 |
| Control Test | | 0.080 | 14:59 |
| Air Blank | | 0.000 | 15:00 |
| Diagnostics Check | | OK | 15:00 |

Data Download Date/Time: 09/11/2019 10:13

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,STEVEN,D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK,STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 08/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 18:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:54

SUBJECT'S LAST NAME: RUPP

DATE/TIME OF TEST: 08/30/2019 19:20:46

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 19:14 |
| Air Blank | | 0.000 | 19:15 |
| Control Test | | 0.080 | 19:15 |
| Air Blank | | 0.000 | 19:15 |
| Subject Sample | # 1 | 0.114 | 19:16 |
| Breath Volume | | 2.226 | 19:16 |
| Air Blank | | 0.000 | 19:17 |
| Air Blank | | 0.000 | 19:18 |
| Subject Sample | # 2 | 0.108 | 19:19 |
| Breath Volume | | 2.394 | 19:19 |
| Air Blank | | 0.000 | 19:19 |
| Control Test | | 0.080 | 19:20 |
| Air Blank | | 0.000 | 19:20 |
| Diagnostics Check | | OK | 19:20 |

Data Download Date/Time: 09/11/2019 10:13

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,MARK,A
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: 5C20,VULETA
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 08/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: SMEDLEY

DATE/TIME OF TEST: 08/31/2019 03:38:44

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 03:35 | |
| Air Blank | | 0.000 | 03:36 | |
| Control Test | | 0.081 | 03:36 | |
| Air Blank | | 0.000 | 03:37 | |
| Subject Sample | # 1 | | 03:37 | Subject Test Refused |
| Breath Volume | | 0.000 | 03:37 | |
| Air Blank | | 0.000 | 03:37 | |
| Control Test | | 0.081 | 03:38 | |
| Air Blank | | 0.000 | 03:38 | |
| Diagnostics Check | | OK | 03:38 | |

Data Download Date/Time: 09/11/2019 10:13

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MAEGAN,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MANGEL,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 08/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 17:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:53

SUBJECT'S LAST NAME: PUGA

DATE/TIME OF TEST: 08/23/2019 18:19:12

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 18:13 |
| Air Blank | | 0.000 | 18:13 |
| Control Test | | 0.079 | 18:14 |
| Air Blank | | 0.000 | 18:14 |
| Subject Sample | # 1 | 0.000 | 18:15 |
| Breath Volume | | 1.503 | 18:15 |
| Air Blank | | 0.000 | 18:15 |
| Air Blank | | 0.000 | 18:17 |
| Subject Sample | # 2 | 0.000 | 18:17 |
| Breath Volume | | 1.152 | 18:17 |
| Air Blank | | 0.000 | 18:18 |
| Control Test | | 0.079 | 18:18 |
| Air Blank | | 0.000 | 18:19 |
| Diagnostics Check | | OK | 18:19 |

Data Download Date/Time: 09/06/2019 05:57

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DEPUTY,.
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: DELEON,DEPUTY
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 08/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:13

SUBJECT'S LAST NAME: BRUN-SOTO

DATE/TIME OF TEST: 08/23/2019 22:10:13

| | | | |
|---|---|---|---|
| Diagnostics Check | OK | 22:09 | |
| Air Blank | | 22:09 | RFI Detect |
| Air Blank | 0.000 | 22:10 | |

Data Download Date/Time: 09/06/2019 05:57

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DEPUTY,.
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: DELEON,DEPUTY
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 08/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:14

SUBJECT'S LAST NAME: BRUN-SOTO

DATE/TIME OF TEST: 08/23/2019 22:25:43

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 22:12 | |
| Air Blank | | 0.000 | 22:12 | |
| Control Test | | 0.080 | 22:12 | |
| Air Blank | | 0.000 | 22:13 | |
| Subject Sample | # 1 | 0.143 | 22:14 | |
| Breath Volume | | 1.343 | 22:14 | |
| Air Blank | | 0.000 | 22:15 | |
| Air Blank | | 0.000 | 22:17 | |
| Subject Sample | # 2 | 0.122 | 22:20 | Volume Not Met |
| Breath Volume | | 0.058 | 22:20 | |
| Air Blank | | 0.000 | 22:20 | |
| Air Blank | | 0.000 | 22:22 | |
| Subject Sample | # 3 | 0.147 | 22:24 | |
| Breath Volume | | 1.558 | 22:24 | |
| Air Blank | | 0.000 | 22:24 | |
| Control Test | | 0.077 | 22:25 | |
| Air Blank | | 0.000 | 22:25 | |
| Diagnostics Check | | OK | 22:25 | |

Data Download Date/Time: 09/06/2019 05:57

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEARSE,ANGIE,D
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: ROLLANDO,T
ARRESTING OFFICER AGENCY: Fellsmere PD
LAST AGENCY INSPECTION: 08/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: KOZELKA

DATE/TIME OF TEST: 08/24/2019 03:47:37

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 03:42 | |
| Air Blank | | 0.000 | 03:42 | |
| Control Test | | 0.079 | 03:43 | |
| Air Blank | | 0.000 | 03:43 | |
| Subject Sample | # 1 | 0.241 | 03:44 | |
| Breath Volume | | 2.386 | 03:44 | |
| Air Blank | | 0.000 | 03:44 | |
| Air Blank | | 0.000 | 03:46 | |
| Subject Sample | # 2 | 0.241 | 03:47 | |
| Breath Volume | | 2.597 | 03:47 | |
| Air Blank | | | 03:47 | RFI Detect |
| Air Blank | | | 03:47 | RFI Detect |

Data Download Date/Time: 09/06/2019 05:57


COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLIDAY,LATONYA,A
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: LEVENAR,T
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 08/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 04:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 08/25/2019 05:54:47

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 05:51 | |
| Air Blank | | 0.000 | 05:51 | |
| Control Test | | 0.079 | 05:51 | |
| Air Blank | | 0.000 | 05:52 | |
| Subject Sample | # 1 | | 05:54 | Sequence Aborted |
| Breath Volume | | 0.000 | 05:54 | |
| Air Blank | | 0.000 | 05:54 | |

Data Download Date/Time: 09/06/2019 05:57

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLIDAY,LATONYA,A
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: LEVENAR,T
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 08/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 04:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:56

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 08/25/2019 06:24:04

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 06:17 |
| Air Blank | | 0.000 | 06:17 |
| Control Test | | 0.079 | 06:18 |
| Air Blank | | 0.000 | 06:18 |
| Subject Sample | # 1 | 0.067 | 06:19 |
| Breath Volume | | 1.257 | 06:19 |
| Air Blank | | 0.000 | 06:20 |
| Air Blank | | 0.000 | 06:22 |
| Subject Sample | # 2 | 0.068 | 06:22 |
| Breath Volume | | 1.726 | 06:22 |
| Air Blank | | 0.000 | 06:23 |
| Control Test | | 0.079 | 06:23 |
| Air Blank | | 0.000 | 06:23 |
| Diagnostics Check | | OK | 06:24 |

Data Download Date/Time: 09/06/2019 05:57

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,DEVIN,C
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: RUIZ,MICHAEL
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 08/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: MCKISSICK

DATE/TIME OF TEST: 08/26/2019 02:08:45

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:00 |
| Air Blank | | 0.000 | 02:01 |
| Control Test | | 0.079 | 02:01 |
| Air Blank | | 0.000 | 02:02 |
| Subject Sample | # 1 | 0.138 | 02:03 |
| Breath Volume | | 1.460 | 02:03 |
| Air Blank | | 0.000 | 02:04 |
| Air Blank | | 0.000 | 02:05 |
| Subject Sample | # 2 | 0.141 | 02:07 |
| Breath Volume | | 1.589 | 02:07 |
| Air Blank | | 0.000 | 02:07 |
| Control Test | | 0.079 | 02:08 |
| Air Blank | | 0.000 | 02:08 |
| Diagnostics Check | | OK | 02:08 |

Data Download Date/Time: 09/06/2019 05:57

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHILE,C,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: FLORES,D
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 08/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: SCHAEFER

DATE/TIME OF TEST: 08/28/2019 23:28:34

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:22 |
| Air Blank | | 0.000 | 23:22 |
| Control Test | | 0.079 | 23:23 |
| Air Blank | | 0.000 | 23:23 |
| Subject Sample | # 1 | 0.000 | 23:24 |
| Breath Volume | | 2.152 | 23:24 |
| Air Blank | | 0.000 | 23:24 |
| Air Blank | | 0.000 | 23:26 |
| Subject Sample | # 2 | 0.000 | 23:27 |
| Breath Volume | | 1.242 | 23:27 |
| Air Blank | | 0.000 | 23:27 |
| Control Test | | 0.079 | 23:27 |
| Air Blank | | 0.000 | 23:28 |
| Diagnostics Check | | OK | 23:28 |

Data Download Date/Time: 09/06/2019 05:57

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAYMOND,FRANK,W
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: BARCARO,BRIANNA
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 08/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: RUBI

DATE/TIME OF TEST: 08/30/2019 01:20:36

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 01:17 | |
| Air Blank | | 0.000 | 01:17 | |
| Control Test | | 0.079 | 01:18 | |
| Air Blank | | 0.000 | 01:18 | |
| Subject Sample | # 1 | | 01:19 | Subject Test Refused |
| Breath Volume | | 0.000 | 01:19 | |
| Air Blank | | 0.000 | 01:19 | |
| Control Test | | 0.080 | 01:20 | |
| Air Blank | | 0.000 | 01:20 | |
| Diagnostics Check | | OK | 01:20 | |

Data Download Date/Time: 09/06/2019 05:57

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FONSECA,GIOVANNY,
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: FONSECA,GIOVANNY
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 08/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 08/30/2019 04:05:30

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:58 |
| Air Blank | | 0.000 | 03:59 |
| Control Test | | 0.079 | 03:59 |
| Air Blank | | 0.000 | 03:59 |
| Subject Sample | # 1 | 0.157 | 04:00 |
| Breath Volume | | 2.367 | 04:00 |
| Air Blank | | 0.000 | 04:01 |
| Air Blank | | 0.000 | 04:03 |
| Subject Sample | # 2 | 0.164 | 04:03 |
| Breath Volume | | 2.199 | 04:03 |
| Air Blank | | 0.000 | 04:04 |
| Control Test | | 0.079 | 04:04 |
| Air Blank | | 0.000 | 04:05 |
| Diagnostics Check | | OK | 04:05 |

Data Download Date/Time: 09/06/2019 05:58

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,TERRENCE,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: JAWORSKI,K
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 08/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 09:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:27

SUBJECT'S LAST NAME: MOLLOY

DATE/TIME OF TEST: 08/30/2019 10:59:21

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 10:52 |
| Air Blank | | 0.000 | 10:52 |
| Control Test | | 0.079 | 10:52 |
| Air Blank | | 0.000 | 10:53 |
| Subject Sample | # 1 | 0.000 | 10:54 |
| Breath Volume | | 2.351 | 10:54 |
| Air Blank | | 0.000 | 10:55 |
| Air Blank | | 0.000 | 10:57 |
| Subject Sample | # 2 | 0.000 | 10:57 |
| Breath Volume | | 2.531 | 10:57 |
| Air Blank | | 0.000 | 10:58 |
| Control Test | | 0.079 | 10:58 |
| Air Blank | | 0.000 | 10:59 |
| Diagnostics Check | | OK | 10:59 |

Data Download Date/Time: 09/06/2019 05:58

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARROQUIN,RICHARD,J
BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)
ARRESTING OFFICER: MARROQUIN,RICHARD
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 08/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: BUI
ARREST TIME: 20:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:42

SUBJECT'S LAST NAME: FENNING

DATE/TIME OF TEST: 08/31/2019 21:32:50

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 21:30 | |
| Air Blank | | 0.000 | 21:30 | |
| Control Test | | 0.079 | 21:30 | |
| Air Blank | | 0.000 | 21:31 | |
| Subject Sample | # 1 | | 21:31 | Subject Test Refused |
| Breath Volume | | 0.000 | 21:31 | |
| Air Blank | | 0.000 | 21:31 | |
| Control Test | | 0.079 | 21:32 | |
| Air Blank | | 0.000 | 21:32 | |
| Diagnostics Check | | OK | 21:32 | |

Data Download Date/Time: 09/06/2019 05:58

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANDY,KEVIN,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: VAVRA,TYLER
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 07/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: GRANT

DATE/TIME OF TEST: 08/01/2019 00:32:47

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:26 |
| Air Blank | | 0.000 | 00:26 |
| Control Test | | 0.083 | 00:26 |
| Air Blank | | 0.000 | 00:27 |
| Subject Sample | # 1 | 0.089 | 00:27 |
| Breath Volume | | 2.750 | 00:27 |
| Air Blank | | 0.000 | 00:28 |
| Air Blank | | 0.000 | 00:30 |
| Subject Sample | # 2 | 0.091 | 00:31 |
| Breath Volume | | 3.148 | 00:31 |
| Air Blank | | 0.000 | 00:31 |
| Control Test | | 0.083 | 00:32 |
| Air Blank | | 0.000 | 00:32 |
| Diagnostics Check | | OK | 00:32 |

Data Download Date/Time: 08/23/2019 05:56

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHILE,C,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: SINGH,T
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 07/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: MULL

DATE/TIME OF TEST: 08/01/2019 22:54:55

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 22:41 | |
| Air Blank | | 0.000 | 22:42 | |
| Control Test | | 0.082 | 22:42 | |
| Air Blank | | 0.000 | 22:43 | |
| Subject Sample | # 1 | 0.160 | 22:46 | Volume Not Met |
| Breath Volume | | 0.000 | 22:46 | |
| Air Blank | | 0.000 | 22:47 | |
| Air Blank | | 0.000 | 22:48 | |
| Subject Sample | # 2 | 0.150 | 22:50 | |
| Breath Volume | | 1.378 | 22:50 | |
| Air Blank | | 0.000 | 22:50 | |
| Air Blank | | 0.000 | 22:52 | |
| Subject Sample | # 3 | 0.148 | 22:53 | |
| Breath Volume | | 1.546 | 22:53 | |
| Air Blank | | 0.000 | 22:54 | |
| Control Test | | 0.080 | 22:54 | |
| Air Blank | | 0.000 | 22:54 | |
| Diagnostics Check | | OK | 22:54 | |

Data Download Date/Time: 08/23/2019 05:56

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,DEVIN,C
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MAJOR,T
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 07/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 08/02/2019 02:29:04

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:21 |
| Air Blank | | 0.000 | 02:21 |
| Control Test | | 0.083 | 02:21 |
| Air Blank | | 0.000 | 02:22 |
| Subject Sample | # 1 | 0.091 | 02:23 |
| Breath Volume | | 1.871 | 02:23 |
| Air Blank | | 0.000 | 02:24 |
| Air Blank | | 0.000 | 02:26 |
| Subject Sample | # 2 | 0.089 | 02:27 |
| Breath Volume | | 1.316 | 02:27 |
| Air Blank | | 0.000 | 02:28 |
| Control Test | | 0.083 | 02:28 |
| Air Blank | | 0.000 | 02:28 |
| Diagnostics Check | | OK | 02:29 |

Data Download Date/Time: 08/23/2019 05:56

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARROQUIN,RICHARD,J
BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)
ARRESTING OFFICER: MARROQUIN,RICHARD
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 07/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: MANN

DATE/TIME OF TEST: 08/03/2019 23:22:53

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:16 |
| Air Blank | | 0.000 | 23:17 |
| Control Test | | 0.083 | 23:17 |
| Air Blank | | 0.000 | 23:18 |
| Subject Sample | # 1 | 0.000 | 23:18 |
| Breath Volume | | 2.902 | 23:18 |
| Air Blank | | 0.000 | 23:19 |
| Air Blank | | 0.000 | 23:21 |
| Subject Sample | # 2 | 0.000 | 23:21 |
| Breath Volume | | 3.113 | 23:21 |
| Air Blank | | 0.000 | 23:21 |
| Control Test | | 0.083 | 23:22 |
| Air Blank | | 0.000 | 23:22 |
| Diagnostics Check | | OK | 23:22 |

Data Download Date/Time: 08/23/2019 05:57

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANDY,KEVIN,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: SANTANA,ROMEO
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 07/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: KENNEDY

DATE/TIME OF TEST: 08/04/2019 03:17:12

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:11 |
| Air Blank | | 0.000 | 03:11 |
| Control Test | | 0.084 | 03:11 |
| Air Blank | | 0.000 | 03:12 |
| Subject Sample | # 1 | 0.000 | 03:12 |
| Breath Volume | | 2.898 | 03:12 |
| Air Blank | | 0.000 | 03:13 |
| Air Blank | | 0.000 | 03:15 |
| Subject Sample | # 2 | 0.000 | 03:15 |
| Breath Volume | | 2.261 | 03:15 |
| Air Blank | | 0.000 | 03:16 |
| Control Test | | 0.083 | 03:16 |
| Air Blank | | 0.000 | 03:17 |
| Diagnostics Check | | OK | 03:17 |

Data Download Date/Time: 08/23/2019 05:57

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OOLEY,JAMES,W
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: DEANTONIO,DANIELA
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 07/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: ANCHORS'

DATE/TIME OF TEST: 08/05/2019 00:04:55

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 23:50 | |
| Air Blank | | 0.000 | 23:51 | |
| Control Test | | 0.083 | 23:51 | |
| Air Blank | | 0.000 | 23:51 | |
| Subject Sample | # 2 | 0.087 | 23:54 | Volume Not Met |
| Breath Volume | | 0.410 | 23:54 | |
| Air Blank | | 0.000 | 23:55 | |
| Air Blank | | 0.000 | 23:57 | |
| Subject Sample | # 3 | 0.101 | 23:57 | |
| Breath Volume | | 1.433 | 23:57 | |
| Air Blank | | 0.000 | 23:58 | |
| Air Blank | | 0.000 | 00:00 | |
| Subject Sample | # 1 | 0.099 | 00:03 | |
| Breath Volume | | 1.265 | 00:03 | |
| Air Blank | | 0.000 | 00:04 | |
| Control Test | | 0.081 | 00:04 | |
| Air Blank | | 0.000 | 00:04 | |
| Diagnostics Check | | OK | 00:04 | |

Data Download Date/Time: 08/23/2019 05:57

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,BRIAN,P
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: HERNANDEZ,DANIEL
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 07/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: HINDS III

DATE/TIME OF TEST: 08/06/2019 01:24:09

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:17 |
| Air Blank | | 0.000 | 01:18 |
| Control Test | | 0.082 | 01:18 |
| Air Blank | | 0.000 | 01:19 |
| Subject Sample | # 1 | 0.000 | 01:19 |
| Breath Volume | | 2.070 | 01:19 |
| Air Blank | | 0.000 | 01:20 |
| Air Blank | | 0.000 | 01:22 |
| Subject Sample | # 2 | 0.000 | 01:22 |
| Breath Volume | | 1.699 | 01:22 |
| Air Blank | | 0.000 | 01:23 |
| Control Test | | 0.083 | 01:23 |
| Air Blank | | 0.000 | 01:24 |
| Diagnostics Check | | OK | 01:24 |

Data Download Date/Time: 08/23/2019 05:57

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FONSECA,GIOVANNY,
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: GREEN,BRIAN
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 07/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: GARCIA DIAZ

DATE/TIME OF TEST: 08/07/2019 03:35:51

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 03:32 | |
| Air Blank | | 0.000 | 03:33 | |
| Control Test | | 0.083 | 03:33 | |
| Air Blank | | 0.000 | 03:34 | |
| Subject Sample | # 1 | | 03:34 | Subject Test Refused |
| Breath Volume | | 0.000 | 03:34 | |
| Air Blank | | 0.000 | 03:34 | |
| Control Test | | 0.083 | 03:35 | |
| Air Blank | | 0.000 | 03:35 | |
| Diagnostics Check | | OK | 03:35 | |

Data Download Date/Time: 08/23/2019 05:57

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIGNORELLI,A,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: FLORES,D
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 07/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 21:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: AURELIA

DATE/TIME OF TEST: 08/07/2019 21:58:44

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 21:49 |
| Air Blank | | 0.000 | 21:50 |
| Control Test | | 0.084 | 21:50 |
| Air Blank | | 0.000 | 21:51 |
| Subject Sample | # 1 | 0.104 | 21:53 |
| Breath Volume | | 1.367 | 21:53 |
| Air Blank | | 0.000 | 21:54 |
| Air Blank | | 0.000 | 21:55 |
| Subject Sample | # 2 | 0.104 | 21:57 |
| Breath Volume | | 1.539 | 21:57 |
| Air Blank | | 0.000 | 21:57 |
| Control Test | | 0.081 | 21:58 |
| Air Blank | | 0.000 | 21:58 |
| Diagnostics Check | | OK | 21:58 |

Data Download Date/Time: 08/23/2019 05:57

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OOLEY,JAMES,W
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: OOLEY,JAMES
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 07/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: SCALFO

DATE/TIME OF TEST: 08/08/2019 03:31:32

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:25 |
| Air Blank | | 0.000 | 03:25 |
| Control Test | | 0.084 | 03:26 |
| Air Blank | | 0.000 | 03:26 |
| Subject Sample | # 1 | 0.000 | 03:27 |
| Breath Volume | | 3.078 | 03:27 |
| Air Blank | | 0.000 | 03:27 |
| Air Blank | | 0.000 | 03:29 |
| Subject Sample | # 2 | 0.000 | 03:30 |
| Breath Volume | | 2.730 | 03:30 |
| Air Blank | | 0.000 | 03:30 |
| Control Test | | 0.083 | 03:30 |
| Air Blank | | 0.000 | 03:31 |
| Diagnostics Check | | OK | 03:31 |

Data Download Date/Time: 08/23/2019 05:57

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HISS,J,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MAJOR,T
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 07/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 19:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: KIFER

DATE/TIME OF TEST: 08/10/2019 21:51:11

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 21:44 |
| Air Blank | | 0.000 | 21:45 |
| Control Test | | 0.082 | 21:45 |
| Air Blank | | 0.000 | 21:45 |
| Subject Sample | # 1 | 0.116 | 21:46 |
| Breath Volume | | 2.531 | 21:46 |
| Air Blank | | 0.000 | 21:46 |
| Air Blank | | 0.000 | 21:48 |
| Subject Sample | # 2 | 0.120 | 21:49 |
| Breath Volume | | 2.066 | 21:49 |
| Air Blank | | 0.000 | 21:50 |
| Control Test | | 0.082 | 21:50 |
| Air Blank | | 0.000 | 21:51 |
| Diagnostics Check | | OK | 21:51 |

Data Download Date/Time: 08/23/2019 05:57

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEARSE,ANGIE,D
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: RUIZ,MICHAEL
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 07/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: LAWSON

DATE/TIME OF TEST: 08/12/2019 01:19:01

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 01:15 | |
| Air Blank | | 0.000 | 01:16 | |
| Control Test | | 0.084 | 01:16 | |
| Air Blank | | 0.000 | 01:16 | |
| Subject Sample | # 1 | | 01:17 | Subject Test Refused |
| Breath Volume | | 0.000 | 01:17 | |
| Air Blank | | 0.000 | 01:18 | |
| Control Test | | 0.084 | 01:18 | |
| Air Blank | | 0.000 | 01:18 | |
| Diagnostics Check | | OK | 01:19 | |

Data Download Date/Time: 08/23/2019 05:57

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OOLEY,JAMES,W
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: OOLEY,JAMES
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 07/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: CONLON

DATE/TIME OF TEST: 08/13/2019 01:53:23

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 01:50 | |
| Air Blank | | 0.000 | 01:51 | |
| Control Test | | 0.083 | 01:51 | |
| Air Blank | | 0.000 | 01:51 | |
| Subject Sample | # 1 | | 01:52 | Subject Test Refused |
| Breath Volume | | 0.000 | 01:52 | |
| Air Blank | | 0.000 | 01:52 | |
| Control Test | | 0.083 | 01:52 | |
| Air Blank | | 0.000 | 01:53 | |
| Diagnostics Check | | OK | 01:53 | |

Data Download Date/Time: 08/23/2019 05:57

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIGNORELLI,A,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: FLORES,D
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 07/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 20:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:44

SUBJECT'S LAST NAME: RUSSELL

DATE/TIME OF TEST: 08/13/2019 21:16:42

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 21:13 | |
| Air Blank | | 0.000 | 21:14 | |
| Control Test | | 0.083 | 21:14 | |
| Air Blank | | 0.000 | 21:15 | |
| Subject Sample | # 1 | | 21:15 | Subject Test Refused |
| Breath Volume | | 0.000 | 21:15 | |
| Air Blank | | 0.000 | 21:15 | |
| Control Test | | 0.083 | 21:16 | |
| Air Blank | | 0.000 | 21:16 | |
| Diagnostics Check | | OK | 21:16 | |

Data Download Date/Time: 08/23/2019 05:57

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIGNORELLI,A,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: RUIZ,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 07/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 21:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 08/13/2019 22:01:17

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 21:58 | |
| Air Blank | | 0.000 | 21:58 | |
| Control Test | | 0.083 | 21:58 | |
| Air Blank | | 0.000 | 21:59 | |
| Subject Sample | # 1 | | 21:59 | Subject Test Refused |
| Breath Volume | | 0.000 | 21:59 | |
| Air Blank | | 0.000 | 22:00 | |
| Control Test | | 0.082 | 22:00 | |
| Air Blank | | 0.000 | 22:01 | |
| Diagnostics Check | | OK | 22:01 | |

Data Download Date/Time: 08/23/2019 05:57

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KANTORSKI,MARK,A
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: FERGUSON,DEPUTY
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 07/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: SCHWEITZER

DATE/TIME OF TEST: 08/14/2019 01:26:22

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 01:23 | |
| Air Blank | | 0.000 | 01:23 | |
| Control Test | | 0.082 | 01:24 | |
| Air Blank | | 0.000 | 01:24 | |
| Subject Sample | # 1 | | 01:24 | Subject Test Refused |
| Breath Volume | | 0.000 | 01:24 | |
| Air Blank | | 0.000 | 01:25 | |
| Control Test | | 0.083 | 01:25 | |
| Air Blank | | 0.000 | 01:26 | |
| Diagnostics Check | | OK | 01:26 | |

Data Download Date/Time: 08/23/2019 05:57

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OOLEY,JAMES,W
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: OOLEY,JAMES
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 07/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: MATHENY

DATE/TIME OF TEST: 08/17/2019 02:32:20

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 02:29 | |
| Air Blank | | 0.000 | 02:29 | |
| Control Test | | 0.083 | 02:30 | |
| Air Blank | | 0.000 | 02:30 | |
| Subject Sample | # 1 | | 02:30 | Subject Test Refused |
| Breath Volume | | 0.000 | 02:30 | |
| Air Blank | | 0.000 | 02:31 | |
| Control Test | | 0.082 | 02:31 | |
| Air Blank | | 0.000 | 02:32 | |
| Diagnostics Check | | OK | 02:32 | |

Data Download Date/Time: 08/23/2019 05:57

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARROQUIN,RICHARD,J
BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)
ARRESTING OFFICER: MARROQUIN,RICHARD
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 07/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: BUI
ARREST TIME: 18:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:49

SUBJECT'S LAST NAME: MAYNARD

DATE/TIME OF TEST: 08/17/2019 20:17:47

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 20:14 | |
| Air Blank | | 0.000 | 20:15 | |
| Control Test | | 0.083 | 20:15 | |
| Air Blank | | 0.000 | 20:16 | |
| Subject Sample | # 1 | | 20:16 | Subject Test Refused |
| Breath Volume | | 0.000 | 20:16 | |
| Air Blank | | 0.000 | 20:16 | |
| Control Test | | 0.083 | 20:17 | |
| Air Blank | | 0.000 | 20:17 | |
| Diagnostics Check | | OK | 20:17 | |

Data Download Date/Time: 08/23/2019 05:57

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OOLEY,JAMES,W
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: OOLEY,JAMES
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 07/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: PARRAMORA

DATE/TIME OF TEST: 08/18/2019 02:41:51

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:34 |
| Air Blank | | 0.000 | 02:35 |
| Control Test | | 0.082 | 02:35 |
| Air Blank | | 0.000 | 02:36 |
| Subject Sample | # 1 | 0.194 | 02:36 |
| Breath Volume | | 1.171 | 02:36 |
| Air Blank | | 0.000 | 02:37 |
| Air Blank | | 0.000 | 02:39 |
| Subject Sample | # 2 | 0.192 | 02:40 |
| Breath Volume | | 1.292 | 02:40 |
| Air Blank | | 0.000 | 02:40 |
| Control Test | | 0.081 | 02:41 |
| Air Blank | | 0.000 | 02:41 |
| Diagnostics Check | | OK | 02:41 |

Data Download Date/Time: 08/23/2019 05:57

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DEPUTY,T
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: 2018,A
ARRESTING OFFICER AGENCY: Fellsmere PD
LAST AGENCY INSPECTION: 07/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: MOJICA

DATE/TIME OF TEST: 08/18/2019 21:46:47

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 21:40 |
| Air Blank | | 0.000 | 21:40 |
| Control Test | | 0.082 | 21:41 |
| Air Blank | | 0.000 | 21:41 |
| Subject Sample | # 1 | 0.058 | 21:42 |
| Breath Volume | | 1.742 | 21:42 |
| Air Blank | | 0.000 | 21:42 |
| Air Blank | | 0.000 | 21:44 |
| Subject Sample | # 2 | 0.058 | 21:45 |
| Breath Volume | | 1.757 | 21:45 |
| Air Blank | | 0.000 | 21:45 |
| Control Test | | 0.082 | 21:46 |
| Air Blank | | 0.000 | 21:46 |
| Diagnostics Check | | OK | 21:46 |

Data Download Date/Time: 08/23/2019 05:57

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OOLEY,JAMES,W
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: OOLEY,JAMES
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 07/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: GIBSON

DATE/TIME OF TEST: 08/19/2019 02:50:41

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:40 |
| Air Blank | | 0.000 | 02:41 |
| Control Test | | 0.082 | 02:41 |
| Air Blank | | 0.000 | 02:42 |
| Subject Sample | # 1 | 0.366 | 02:42 |
| Breath Volume | | 2.496 | 02:42 |
| Air Blank | | 0.000 | 02:43 |
| Air Blank | | 0.000 | 02:45 |
| Subject Sample | # 2 | 0.339 | 02:45 |
| Breath Volume | | 2.042 | 02:45 |
| Air Blank | | 0.000 | 02:46 |
| Air Blank | | 0.000 | 02:48 |
| Subject Sample | # 3 | 0.360 | 02:48 |
| Breath Volume | | 2.468 | 02:48 |
| Air Blank | | 0.000 | 02:49 |
| Control Test | | 0.078 | 02:50 |
| Air Blank | | 0.000 | 02:50 |
| Diagnostics Check | | OK | 02:50 |

Data Download Date/Time: 08/23/2019 05:57

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MAEGAN,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: VAVRA,T
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 07/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: MANGIAPANI

DATE/TIME OF TEST: 08/19/2019 23:25:10

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:18 |
| Air Blank | | 0.000 | 23:19 |
| Control Test | | 0.083 | 23:19 |
| Air Blank | | 0.000 | 23:19 |
| Subject Sample | # 1 | 0.186 | 23:20 |
| Breath Volume | | 1.531 | 23:20 |
| Air Blank | | 0.000 | 23:21 |
| Air Blank | | 0.000 | 23:23 |
| Subject Sample | # 2 | 0.195 | 23:23 |
| Breath Volume | | 1.207 | 23:23 |
| Air Blank | | 0.000 | 23:24 |
| Control Test | | 0.081 | 23:24 |
| Air Blank | | 0.000 | 23:25 |
| Diagnostics Check | | OK | 23:25 |

Data Download Date/Time: 08/23/2019 05:57

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MAEGAN,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: QUINN,T
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 07/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: PERLSTEIN

DATE/TIME OF TEST: 08/20/2019 23:26:17

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:16 |
| Air Blank | | 0.000 | 23:17 |
| Control Test | | 0.082 | 23:17 |
| Air Blank | | 0.000 | 23:18 |
| Subject Sample | # 1 | 0.198 | 23:21 |
| Breath Volume | | 1.414 | 23:21 |
| Air Blank | | 0.000 | 23:21 |
| Air Blank | | 0.000 | 23:23 |
| Subject Sample | # 2 | 0.204 | 23:24 |
| Breath Volume | | 1.238 | 23:24 |
| Air Blank | | 0.000 | 23:25 |
| Control Test | | 0.081 | 23:25 |
| Air Blank | | 0.000 | 23:26 |
| Diagnostics Check | | OK | 23:26 |

Data Download Date/Time: 08/23/2019 05:57

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHILE,C,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MANGEL,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 07/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:01

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 08/22/2019 23:39:32

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:32 |
| Air Blank | | 0.000 | 23:33 |
| Control Test | | 0.082 | 23:33 |
| Air Blank | | 0.000 | 23:33 |
| Subject Sample | # 1 | 0.297 | 23:34 |
| Breath Volume | | 1.667 | 23:34 |
| Air Blank | | 0.000 | 23:35 |
| Air Blank | | 0.000 | 23:37 |
| Subject Sample | # 2 | 0.299 | 23:37 |
| Breath Volume | | 1.386 | 23:37 |
| Air Blank | | 0.000 | 23:38 |
| Control Test | | 0.081 | 23:38 |
| Air Blank | | 0.000 | 23:39 |
| Diagnostics Check | | OK | 23:39 |

Data Download Date/Time: 08/23/2019 05:57

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VEGA,JASON,
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: GALVIN,MIGUEL
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 07/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 07:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:43

SUBJECT'S LAST NAME: FICKLER

DATE/TIME OF TEST: 08/09/2019 08:17:40

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 08:10 | |
| Air Blank | | 0.000 | 08:10 | |
| Control Test | | 0.080 | 08:11 | |
| Air Blank | | 0.000 | 08:11 | |
| Subject Sample | # 1 | 0.217 | 08:13 | |
| Breath Volume | | 1.664 | 08:13 | |
| Air Blank | | 0.000 | 08:14 | |
| Air Blank | | 0.000 | 08:16 | |
| Subject Sample | # 2 | 0.227 | 08:17 | |
| Breath Volume | | 1.394 | 08:17 | |
| Air Blank | | | 08:17 | Improper Sample |
| Air Blank | | 0.000 | 08:17 | |

Data Download Date/Time: 08/29/2019 09:51

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VEGA,JASON,
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: BAKER,DUSTIN
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 07/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 08/15/2019 17:24:41

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 17:15 |
| Air Blank | | 0.000 | 17:15 |
| Control Test | | 0.081 | 17:15 |
| Air Blank | | 0.000 | 17:16 |
| Subject Sample | # 1 | 0.298 | 17:19 |
| Breath Volume | | 1.500 | 17:19 |
| Air Blank | | 0.000 | 17:19 |
| Air Blank | | 0.000 | 17:21 |
| Subject Sample | # 2 | 0.282 | 17:23 |
| Breath Volume | | 1.367 | 17:23 |
| Air Blank | | 0.000 | 17:23 |
| Control Test | | 0.080 | 17:24 |
| Air Blank | | 0.000 | 17:24 |
| Diagnostics Check | | OK | 17:24 |

Data Download Date/Time: 08/29/2019 09:51

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FILIPOWICZ,KEVIN,M
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: STRONG,JONATHAN
ARRESTING OFFICER AGENCY: Arcadia PD
LAST AGENCY INSPECTION: 07/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:18

SUBJECT'S LAST NAME: SANTOS

DATE/TIME OF TEST: 08/18/2019 19:46:02

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 19:38 |
| Air Blank | | 0.000 | 19:39 |
| Control Test | | 0.081 | 19:39 |
| Air Blank | | 0.000 | 19:39 |
| Subject Sample | # 1 | 0.320 | 19:41 |
| Breath Volume | | 1.894 | 19:41 |
| Air Blank | | 0.000 | 19:41 |
| Air Blank | | 0.000 | 19:43 |
| Subject Sample | # 2 | 0.330 | 19:44 |
| Breath Volume | | 2.140 | 19:44 |
| Air Blank | | 0.000 | 19:45 |
| Control Test | | 0.081 | 19:45 |
| Air Blank | | 0.000 | 19:45 |
| Diagnostics Check | | OK | 19:46 |

Data Download Date/Time: 08/29/2019 09:51

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DITTMAN,LAVERNE,T
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: COX,RILEY
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 07/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 08/19/2019 20:58:09

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 20:50 |
| Air Blank | | 0.000 | 20:51 |
| Control Test | | 0.080 | 20:51 |
| Air Blank | | 0.000 | 20:52 |
| Subject Sample | # 1 | 0.164 | 20:53 |
| Breath Volume | | 1.160 | 20:53 |
| Air Blank | | 0.000 | 20:53 |
| Air Blank | | 0.000 | 20:55 |
| Subject Sample | # 2 | 0.171 | 20:56 |
| Breath Volume | | 1.531 | 20:56 |
| Air Blank | | 0.000 | 20:57 |
| Control Test | | 0.080 | 20:57 |
| Air Blank | | 0.000 | 20:58 |
| Diagnostics Check | | OK | 20:58 |

Data Download Date/Time: 08/29/2019 09:51

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,JOHN,P
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: BUENRESTRO,PEDRO
ARRESTING OFFICER AGENCY: Arcadia PD
LAST AGENCY INSPECTION: 07/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:10

SUBJECT'S LAST NAME: PRODY

DATE/TIME OF TEST: 08/27/2019 17:04:18

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 16:50 |
| Air Blank | | 0.000 | 16:51 |
| Control Test | | 0.080 | 16:51 |
| Air Blank | | 0.000 | 16:51 |
| Subject Sample | # 1 | 0.329 | 16:55 |
| Breath Volume | | 1.753 | 16:55 |
| Air Blank | | 0.000 | 16:55 |
| Air Blank | | 0.000 | 16:57 |
| Subject Sample | # 2 | 0.355 | 16:59 |
| Breath Volume | | 1.804 | 16:59 |
| Air Blank | | 0.000 | 17:00 |
| Air Blank | | 0.000 | 17:02 |
| Subject Sample | # 3 | 0.359 | 17:02 |
| Breath Volume | | 2.679 | 17:02 |
| Air Blank | | 0.000 | 17:03 |
| Control Test | | 0.079 | 17:03 |
| Air Blank | | 0.000 | 17:04 |
| Diagnostics Check | | OK | 17:04 |

Data Download Date/Time: 08/29/2019 09:51

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HACKETT',SEAN,J
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: MURCHINSON,RONNIE
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 08/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 08:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:07

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 08/17/2019 09:50:29

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 09:37 | |
| Air Blank | | 0.000 | 09:37 | |
| Control Test | | 0.079 | 09:37 | |
| Air Blank | | 0.000 | 09:38 | |
| Subject Sample | # 1 | 0.000 | 09:41 | Volume Not Met |
| Breath Volume | | 0.578 | 09:41 | |
| Air Blank | | 0.000 | 09:41 | |
| Air Blank | | 0.000 | 09:43 | |
| Subject Sample | # 2 | 0.000 | 09:44 | |
| Breath Volume | | 2.863 | 09:44 | |
| Air Blank | | 0.000 | 09:44 | |
| Air Blank | | 0.000 | 09:46 | |
| Subject Sample | # 3 | 0.000 | 09:49 | |
| Breath Volume | | 1.828 | 09:49 | |
| Air Blank | | 0.000 | 09:49 | |
| Control Test | | 0.079 | 09:49 | |
| Air Blank | | 0.000 | 09:50 | |
| Diagnostics Check | | OK | 09:50 | |

Data Download Date/Time: 09/22/2019 08:39

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIEVES,MATTHEW,J
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: PINSON,NICHOLAS
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 08/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 08/28/2019 23:53:00

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 23:49 | |
| Air Blank | | 0.000 | 23:49 | |
| Control Test | | 0.080 | 23:50 | |
| Air Blank | | 0.000 | 23:50 | |
| Subject Sample | # 1 | | 23:51 | Subject Test Refused |
| Breath Volume | | 0.000 | 23:51 | |
| Air Blank | | 0.000 | 23:52 | |
| Control Test | | 0.081 | 23:52 | |
| Air Blank | | 0.000 | 23:52 | |
| Diagnostics Check | | OK | 23:53 | |

Data Download Date/Time: 09/22/2019 08:39

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUMANN,DEP,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: CANTWELL,TRP
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: TREJO

DATE/TIME OF TEST: 08/11/2019 02:31:14

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:25 |
| Air Blank | | 0.000 | 02:25 |
| Control Test | | 0.079 | 02:25 |
| Air Blank | | 0.000 | 02:26 |
| Subject Sample | # 1 | 0.167 | 02:26 |
| Breath Volume | | 2.695 | 02:26 |
| Air Blank | | 0.000 | 02:27 |
| Air Blank | | 0.000 | 02:29 |
| Subject Sample | # 2 | 0.172 | 02:29 |
| Breath Volume | | 2.472 | 02:29 |
| Air Blank | | 0.000 | 02:30 |
| Control Test | | 0.079 | 02:30 |
| Air Blank | | 0.000 | 02:31 |
| Diagnostics Check | | OK | 02:31 |

Data Download Date/Time: 08/19/2019 17:24

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUMANN,DEP,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SCHUMANN,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: MEINKET

DATE/TIME OF TEST: 08/11/2019 23:17:38

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:10 |
| Air Blank | | 0.000 | 23:11 |
| Control Test | | 0.078 | 23:11 |
| Air Blank | | 0.000 | 23:12 |
| Subject Sample | # 1 | 0.169 | 23:12 |
| Breath Volume | | 1.488 | 23:12 |
| Air Blank | | 0.000 | 23:13 |
| Air Blank | | 0.000 | 23:14 |
| Subject Sample | # 2 | 0.166 | 23:16 |
| Breath Volume | | 2.027 | 23:16 |
| Air Blank | | 0.000 | 23:16 |
| Control Test | | 0.078 | 23:17 |
| Air Blank | | 0.000 | 23:17 |
| Diagnostics Check | | OK | 23:17 |

Data Download Date/Time: 08/19/2019 17:24

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUMANN,DEP,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: CANTWELL,TRP
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: GUSTIN

DATE/TIME OF TEST: 08/14/2019 23:57:35

| | | | |
|---|---|---|---|
| Diagnostics Check | OK | 23:56 | |
| Air Blank | 0.000 | 23:56 | |
| Control Test | 0.079 | 23:56 | |
| Air Blank | | 23:57 | RFI Detect |
| Air Blank | 0.000 | 23:57 | |

Data Download Date/Time: 08/19/2019 17:24

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUMANN,DEP,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: CANTWELL,TRP
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: GUSTIN

DATE/TIME OF TEST: 08/15/2019 00:04:59

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 23:58 | |
| Air Blank | | 0.000 | 23:59 | |
| Control Test | | 0.079 | 23:59 | |
| Air Blank | | 0.000 | 00:00 | |
| Subject Sample | # 1 | 0.266 | 00:00 | |
| Breath Volume | | 1.121 | 00:00 | |
| Air Blank | | 0.000 | 00:01 | |
| Air Blank | | 0.000 | 00:03 | |
| Subject Sample | # 2 | | 00:03 | Slope Not Met |
| Breath Volume | | 0.910 | 00:03 | |
| Air Blank | | 0.000 | 00:04 | |
| Control Test | | 0.078 | 00:04 | |
| Air Blank | | 0.000 | 00:04 | |
| Diagnostics Check | | OK | 00:04 | |

Data Download Date/Time: 08/19/2019 17:24

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUMANN,DEP,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: CANTWELL,TRP
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: GUSTIN

DATE/TIME OF TEST: 08/15/2019 00:06:48

| | | | |
|---|---|---|---|
| Diagnostics Check | OK | 00:06 | |
| Air Blank | | 00:06 | RFI Detect |
| Air Blank | | 00:06 | RFI Detect |

Data Download Date/Time: 08/19/2019 17:24


COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUMANN,DEP,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: CANTWELL,TRP
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: LAMBERT

DATE/TIME OF TEST: 08/16/2019 02:05:46

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:59 |
| Air Blank | | 0.000 | 02:00 |
| Control Test | | 0.079 | 02:00 |
| Air Blank | | 0.000 | 02:00 |
| Subject Sample | # 1 | 0.182 | 02:01 |
| Breath Volume | | 3.253 | 02:01 |
| Air Blank | | 0.000 | 02:01 |
| Air Blank | | 0.000 | 02:03 |
| Subject Sample | # 2 | 0.174 | 02:04 |
| Breath Volume | | 3.613 | 02:04 |
| Air Blank | | 0.000 | 02:04 |
| Control Test | | 0.078 | 02:05 |
| Air Blank | | 0.000 | 02:05 |
| Diagnostics Check | | OK | 02:05 |

Data Download Date/Time: 08/19/2019 17:24

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WATSON,D,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: WATSON,D
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: HELMUTH

DATE/TIME OF TEST: 08/01/2019 02:06:26

| | | | |
|---|---|---|---|
| Diagnostics Check | OK | 02:05 | |
| Air Blank | 0.000 | 02:05 | |
| Control Test | 0.079 | 02:05 | |
| Air Blank | | 02:06 | RFI Detect |
| Air Blank | | 02:06 | RFI Detect |

Data Download Date/Time: 08/19/2019 16:30


COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WATSON,D,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: WATSON,D
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: HELMUTH

DATE/TIME OF TEST: 08/01/2019 02:12:19

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 02:08 | |
| Air Blank | | 0.000 | 02:09 | |
| Control Test | | 0.079 | 02:09 | |
| Air Blank | | 0.000 | 02:09 | |
| Subject Sample | # 1 | 0.101 | 02:10 | |
| Breath Volume | | 1.875 | 02:10 | |
| Air Blank | | | 02:10 | RFI Detect |
| Air Blank | | 0.000 | 02:11 | |
| Control Test | | 0.080 | 02:11 | |
| Air Blank | | | 02:11 | RFI Detect |
| Air Blank | | 0.000 | 02:12 | |

Data Download Date/Time: 08/19/2019 16:30

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WATSON,D,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: WATSON,D
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: HELMUTH

DATE/TIME OF TEST: 08/01/2019 02:20:44

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:14 |
| Air Blank | | 0.000 | 02:14 |
| Control Test | | 0.078 | 02:14 |
| Air Blank | | 0.000 | 02:15 |
| Subject Sample | # 1 | 0.111 | 02:16 |
| Breath Volume | | 1.738 | 02:16 |
| Air Blank | | 0.000 | 02:16 |
| Air Blank | | 0.000 | 02:18 |
| Subject Sample | # 2 | 0.110 | 02:19 |
| Breath Volume | | 1.472 | 02:19 |
| Air Blank | | 0.000 | 02:19 |
| Control Test | | 0.079 | 02:20 |
| Air Blank | | 0.000 | 02:20 |
| Diagnostics Check | | OK | 02:20 |

Data Download Date/Time: 08/19/2019 16:30

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSTYAL,MATT,A
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KOSTYAL,MATTHEW
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: DRECHSLER

DATE/TIME OF TEST: 08/01/2019 21:56:52

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 21:50 |
| Air Blank | | 0.000 | 21:50 |
| Control Test | | 0.080 | 21:50 |
| Air Blank | | 0.000 | 21:51 |
| Subject Sample | # 1 | 0.021 | 21:51 |
| Breath Volume | | 1.960 | 21:51 |
| Air Blank | | 0.000 | 21:52 |
| Air Blank | | 0.000 | 21:54 |
| Subject Sample | # 2 | 0.019 | 21:55 |
| Breath Volume | | 1.179 | 21:55 |
| Air Blank | | 0.000 | 21:55 |
| Control Test | | 0.080 | 21:56 |
| Air Blank | | 0.000 | 21:56 |
| Diagnostics Check | | OK | 21:56 |

Data Download Date/Time: 08/19/2019 16:30

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAMORS,JASON,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: CONTORNO,OFC
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: CONYERS

DATE/TIME OF TEST: 08/02/2019 02:43:40

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:36 |
| Air Blank | | 0.000 | 02:37 |
| Control Test | | 0.079 | 02:37 |
| Air Blank | | 0.000 | 02:37 |
| Subject Sample | # 1 | 0.220 | 02:38 |
| Breath Volume | | 2.523 | 02:38 |
| Air Blank | | 0.000 | 02:39 |
| Air Blank | | 0.000 | 02:41 |
| Subject Sample | # 2 | 0.201 | 02:42 |
| Breath Volume | | 1.464 | 02:42 |
| Air Blank | | 0.000 | 02:42 |
| Control Test | | 0.079 | 02:43 |
| Air Blank | | 0.000 | 02:43 |
| Diagnostics Check | | OK | 02:43 |

Data Download Date/Time: 08/19/2019 16:30

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLINT,SERGEANT,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FLINT,SERGEANT
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: MCDONOUGH

DATE/TIME OF TEST: 08/03/2019 00:55:09

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 00:43 | |
| Air Blank | | 0.000 | 00:44 | |
| Control Test | | 0.079 | 00:44 | |
| Air Blank | | 0.000 | 00:45 | |
| Subject Sample | # 1 | | 00:48 | No Sample Provided |
| Breath Volume | | 0.000 | 00:48 | |
| Air Blank | | 0.000 | 00:48 | |
| Air Blank | | 0.000 | 00:50 | |
| Subject Sample | # 2 | | 00:53 | No Sample Provided |
| Breath Volume | | 0.000 | 00:53 | |
| Air Blank | | 0.000 | 00:54 | |
| Control Test | | 0.080 | 00:54 | |
| Air Blank | | 0.000 | 00:55 | |
| Diagnostics Check | | OK | 00:55 | |

Data Download Date/Time: 08/19/2019 16:30

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLINT,SERGEANT,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FLINT,SERGEANT
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: MCDONOUGH

DATE/TIME OF TEST: 08/03/2019 01:15:13

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:07 |
| Air Blank | | 0.000 | 01:07 |
| Control Test | | 0.079 | 01:07 |
| Air Blank | | 0.000 | 01:08 |
| Subject Sample | # 1 | 0.099 | 01:10 |
| Breath Volume | | 2.289 | 01:10 |
| Air Blank | | 0.000 | 01:11 |
| Air Blank | | 0.000 | 01:13 |
| Subject Sample | # 2 | 0.100 | 01:13 |
| Breath Volume | | 2.363 | 01:13 |
| Air Blank | | 0.000 | 01:14 |
| Control Test | | 0.079 | 01:14 |
| Air Blank | | 0.000 | 01:15 |
| Diagnostics Check | | OK | 01:15 |

Data Download Date/Time: 08/19/2019 16:30

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,NICKKI,N
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: OSTROSKI,N
ARRESTING OFFICER AGENCY: Venice PD
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: WISEGERHOF

DATE/TIME OF TEST: 08/03/2019 01:54:24

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:47 |
| Air Blank | | 0.000 | 01:47 |
| Control Test | | 0.080 | 01:47 |
| Air Blank | | 0.000 | 01:48 |
| Subject Sample | # 1 | 0.125 | 01:49 |
| Breath Volume | | 3.144 | 01:49 |
| Air Blank | | 0.000 | 01:50 |
| Air Blank | | 0.000 | 01:51 |
| Subject Sample | # 2 | 0.122 | 01:52 |
| Breath Volume | | 2.605 | 01:52 |
| Air Blank | | 0.000 | 01:53 |
| Control Test | | 0.079 | 01:53 |
| Air Blank | | 0.000 | 01:54 |
| Diagnostics Check | | OK | 01:54 |

Data Download Date/Time: 08/19/2019 16:30

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACOSTA-BURGOS,KASSANDRA,J
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ACOSTA-BURGOS,KASSANDRA
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 05:30
GENDER: Male

OBSERVATION PERIOD BEGAN: 08:00

SUBJECT'S LAST NAME: SCHWIND

DATE/TIME OF TEST: 08/03/2019 08:26:59

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 08:20 |
| Air Blank | | 0.000 | 08:21 |
| Control Test | | 0.079 | 08:21 |
| Air Blank | | 0.000 | 08:21 |
| Subject Sample | # 1 | 0.111 | 08:22 |
| Breath Volume | | 2.765 | 08:22 |
| Air Blank | | 0.000 | 08:23 |
| Air Blank | | 0.000 | 08:25 |
| Subject Sample | # 2 | 0.111 | 08:25 |
| Breath Volume | | 2.863 | 08:25 |
| Air Blank | | 0.000 | 08:26 |
| Control Test | | 0.078 | 08:26 |
| Air Blank | | 0.000 | 08:26 |
| Diagnostics Check | | OK | 08:26 |

Data Download Date/Time: 08/19/2019 16:30

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,NICKKI,N
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BALES,TIM
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: HEINTZ

DATE/TIME OF TEST: 08/04/2019 00:08:24

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 00:04 | |
| Air Blank | | 0.000 | 00:04 | |
| Control Test | | 0.079 | 00:04 | |
| Air Blank | | 0.000 | 00:05 | |
| Subject Sample | # 1 | | 00:07 | Subject Test Refused |
| Breath Volume | | 0.000 | 00:07 | |
| Air Blank | | 0.000 | 00:07 | |
| Control Test | | 0.079 | 00:07 | |
| Air Blank | | 0.000 | 00:08 | |
| Diagnostics Check | | OK | 00:08 | |

Data Download Date/Time: 08/19/2019 16:30

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,NICKKI,N
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BALES,TIM
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: HEINTZ

DATE/TIME OF TEST: 08/04/2019 00:18:55

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:11 |
| Air Blank | | 0.000 | 00:12 |
| Control Test | | 0.080 | 00:12 |
| Air Blank | | 0.000 | 00:13 |
| Subject Sample | # 1 | 0.097 | 00:13 |
| Breath Volume | | 1.164 | 00:13 |
| Air Blank | | 0.000 | 00:14 |
| Air Blank | | 0.000 | 00:16 |
| Subject Sample | # 2 | 0.102 | 00:17 |
| Breath Volume | | 2.847 | 00:17 |
| Air Blank | | 0.000 | 00:18 |
| Control Test | | 0.078 | 00:18 |
| Air Blank | | 0.000 | 00:18 |
| Diagnostics Check | | OK | 00:18 |

Data Download Date/Time: 08/19/2019 16:30

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,NICKKI,N
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BUTLER,C
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:08

SUBJECT'S LAST NAME: PEOPLES

DATE/TIME OF TEST: 08/04/2019 21:03:51

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 20:57 |
| Air Blank | | 0.000 | 20:58 |
| Control Test | | 0.079 | 20:58 |
| Air Blank | | 0.000 | 20:58 |
| Subject Sample | # 1 | 0.157 | 20:59 |
| Breath Volume | | 2.261 | 20:59 |
| Air Blank | | 0.000 | 21:00 |
| Air Blank | | 0.000 | 21:01 |
| Subject Sample | # 2 | 0.151 | 21:02 |
| Breath Volume | | 1.132 | 21:02 |
| Air Blank | | 0.000 | 21:02 |
| Control Test | | 0.079 | 21:03 |
| Air Blank | | 0.000 | 21:03 |
| Diagnostics Check | | OK | 21:03 |

Data Download Date/Time: 08/19/2019 16:30

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSTYAL,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KOSTYAL,MATTHEW
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: LINDER

DATE/TIME OF TEST: 08/06/2019 02:13:04

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:06 |
| Air Blank | | 0.000 | 02:07 |
| Control Test | | 0.079 | 02:07 |
| Air Blank | | 0.000 | 02:08 |
| Subject Sample | # 1 | 0.138 | 02:08 |
| Breath Volume | | 4.753 | 02:08 |
| Air Blank | | 0.000 | 02:09 |
| Air Blank | | 0.000 | 02:11 |
| Subject Sample | # 2 | 0.135 | 02:11 |
| Breath Volume | | 4.160 | 02:11 |
| Air Blank | | 0.000 | 02:12 |
| Control Test | | 0.079 | 02:12 |
| Air Blank | | 0.000 | 02:12 |
| Diagnostics Check | | OK | 02:13 |

Data Download Date/Time: 08/19/2019 16:30

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WATSON,D,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: WATSON,D
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: GLASZ

DATE/TIME OF TEST: 08/07/2019 04:41:08

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:34 |
| Air Blank | | 0.000 | 04:35 |
| Control Test | | 0.078 | 04:35 |
| Air Blank | | 0.000 | 04:36 |
| Subject Sample | # 1 | 0.148 | 04:36 |
| Breath Volume | | 1.929 | 04:36 |
| Air Blank | | 0.000 | 04:37 |
| Air Blank | | 0.000 | 04:39 |
| Subject Sample | # 2 | 0.140 | 04:39 |
| Breath Volume | | 1.449 | 04:39 |
| Air Blank | | 0.000 | 04:40 |
| Control Test | | 0.079 | 04:40 |
| Air Blank | | 0.000 | 04:41 |
| Diagnostics Check | | OK | 04:41 |

Data Download Date/Time: 08/19/2019 16:30

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,J,S
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BARR,J
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: TRANIER

DATE/TIME OF TEST: 08/09/2019 03:38:19

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:31 |
| Air Blank | | 0.000 | 03:32 |
| Control Test | | 0.080 | 03:32 |
| Air Blank | | 0.000 | 03:33 |
| Subject Sample | # 1 | 0.109 | 03:33 |
| Breath Volume | | 2.210 | 03:33 |
| Air Blank | | 0.000 | 03:34 |
| Air Blank | | 0.000 | 03:36 |
| Subject Sample | # 2 | 0.110 | 03:36 |
| Breath Volume | | 2.722 | 03:36 |
| Air Blank | | 0.000 | 03:37 |
| Control Test | | 0.080 | 03:37 |
| Air Blank | | 0.000 | 03:38 |
| Diagnostics Check | | OK | 03:38 |

Data Download Date/Time: 08/19/2019 16:30

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,SHAUN,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MARTIN,SHAUN
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: MCFADDEN

DATE/TIME OF TEST: 08/10/2019 03:19:12

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:12 |
| Air Blank | | 0.000 | 03:13 |
| Control Test | | 0.079 | 03:13 |
| Air Blank | | 0.000 | 03:14 |
| Subject Sample | # 1 | 0.139 | 03:14 |
| Breath Volume | | 2.671 | 03:14 |
| Air Blank | | 0.000 | 03:15 |
| Air Blank | | 0.000 | 03:17 |
| Subject Sample | # 2 | 0.134 | 03:17 |
| Breath Volume | | 2.781 | 03:17 |
| Air Blank | | 0.000 | 03:18 |
| Control Test | | 0.077 | 03:18 |
| Air Blank | | 0.000 | 03:19 |
| Diagnostics Check | | OK | 03:19 |

Data Download Date/Time: 08/19/2019 16:30

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WATSON,D,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: WATSON,D
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: SIMON

DATE/TIME OF TEST: 08/10/2019 21:55:33

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 21:48 |
| Air Blank | | 0.000 | 21:49 |
| Control Test | | 0.079 | 21:49 |
| Air Blank | | 0.000 | 21:50 |
| Subject Sample | # 1 | 0.115 | 21:51 |
| Breath Volume | | 1.324 | 21:51 |
| Air Blank | | 0.000 | 21:51 |
| Air Blank | | 0.000 | 21:53 |
| Subject Sample | # 2 | 0.114 | 21:54 |
| Breath Volume | | 1.242 | 21:54 |
| Air Blank | | 0.000 | 21:54 |
| Control Test | | 0.079 | 21:54 |
| Air Blank | | 0.000 | 21:55 |
| Diagnostics Check | | OK | 21:55 |

Data Download Date/Time: 08/19/2019 16:30

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLYNN,MICHAEL,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MILLER,R
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: HAMMOND

DATE/TIME OF TEST: 08/11/2019 03:05:08

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:58 |
| Air Blank | | 0.000 | 02:58 |
| Control Test | | 0.079 | 02:59 |
| Air Blank | | 0.000 | 02:59 |
| Subject Sample | # 1 | 0.194 | 03:00 |
| Breath Volume | | 1.695 | 03:00 |
| Air Blank | | 0.000 | 03:01 |
| Air Blank | | 0.000 | 03:03 |
| Subject Sample | # 2 | 0.196 | 03:03 |
| Breath Volume | | 2.214 | 03:03 |
| Air Blank | | 0.000 | 03:04 |
| Control Test | | 0.080 | 03:04 |
| Air Blank | | 0.000 | 03:05 |
| Diagnostics Check | | OK | 03:05 |

Data Download Date/Time: 08/19/2019 16:30

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERKINS,OFC,W
BREATH TEST OPERATOR AGENCY: Sarasota PD
ARRESTING OFFICER: BALES,OFC
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: BATTIS

DATE/TIME OF TEST: 08/11/2019 03:17:42

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:10 |
| Air Blank | | 0.000 | 03:10 |
| Control Test | | 0.080 | 03:11 |
| Air Blank | | 0.000 | 03:11 |
| Subject Sample | # 1 | 0.167 | 03:12 |
| Breath Volume | | 1.679 | 03:12 |
| Air Blank | | 0.000 | 03:13 |
| Air Blank | | 0.000 | 03:15 |
| Subject Sample | # 2 | 0.166 | 03:16 |
| Breath Volume | | 1.328 | 03:16 |
| Air Blank | | 0.000 | 03:16 |
| Control Test | | 0.078 | 03:17 |
| Air Blank | | 0.000 | 03:17 |
| Diagnostics Check | | OK | 03:17 |

Data Download Date/Time: 08/19/2019 16:30

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,SHAUN,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MARTIN,SHAUN
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:24

SUBJECT'S LAST NAME: MANOLO

DATE/TIME OF TEST: 08/11/2019 21:58:18

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 21:51 |
| Air Blank | | 0.000 | 21:52 |
| Control Test | | 0.080 | 21:52 |
| Air Blank | | 0.000 | 21:52 |
| Subject Sample | # 1 | 0.178 | 21:53 |
| Breath Volume | | 2.746 | 21:53 |
| Air Blank | | 0.000 | 21:54 |
| Air Blank | | 0.000 | 21:56 |
| Subject Sample | # 2 | 0.166 | 21:56 |
| Breath Volume | | 2.550 | 21:56 |
| Air Blank | | 0.000 | 21:57 |
| Control Test | | 0.080 | 21:57 |
| Air Blank | | 0.000 | 21:58 |
| Diagnostics Check | | OK | 21:58 |

Data Download Date/Time: 08/19/2019 16:30

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAMORS,JASON,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MILLER,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: TOEPEL

DATE/TIME OF TEST: 08/11/2019 23:05:11

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 22:57 |
| Air Blank | | 0.000 | 22:58 |
| Control Test | | 0.080 | 22:58 |
| Air Blank | | 0.000 | 22:58 |
| Subject Sample | # 1 | 0.136 | 22:59 |
| Breath Volume | | 1.175 | 22:59 |
| Air Blank | | 0.000 | 23:00 |
| Air Blank | | 0.000 | 23:02 |
| Subject Sample | # 2 | 0.137 | 23:03 |
| Breath Volume | | 1.664 | 23:03 |
| Air Blank | | 0.000 | 23:04 |
| Control Test | | 0.080 | 23:04 |
| Air Blank | | 0.000 | 23:05 |
| Diagnostics Check | | OK | 23:05 |

Data Download Date/Time: 08/19/2019 16:30

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,J,S
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BARR,J
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: CARMINE

DATE/TIME OF TEST: 08/13/2019 01:32:45

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:26 |
| Air Blank | | 0.000 | 01:27 |
| Control Test | | 0.079 | 01:27 |
| Air Blank | | 0.000 | 01:27 |
| Subject Sample | # 1 | 0.180 | 01:28 |
| Breath Volume | | 2.523 | 01:28 |
| Air Blank | | 0.000 | 01:28 |
| Air Blank | | 0.000 | 01:30 |
| Subject Sample | # 2 | 0.170 | 01:31 |
| Breath Volume | | 3.613 | 01:31 |
| Air Blank | | 0.000 | 01:31 |
| Control Test | | 0.079 | 01:32 |
| Air Blank | | 0.000 | 01:32 |
| Diagnostics Check | | OK | 01:32 |

Data Download Date/Time: 08/19/2019 16:30

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,SHAUN,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MARTIN,SHAUN
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: LYNCH

DATE/TIME OF TEST: 08/15/2019 00:34:18

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:27 |
| Air Blank | | 0.000 | 00:27 |
| Control Test | | 0.079 | 00:28 |
| Air Blank | | 0.000 | 00:28 |
| Subject Sample | # 1 | 0.166 | 00:29 |
| Breath Volume | | 3.875 | 00:29 |
| Air Blank | | 0.000 | 00:30 |
| Air Blank | | 0.000 | 00:32 |
| Subject Sample | # 2 | 0.177 | 00:32 |
| Breath Volume | | 3.570 | 00:32 |
| Air Blank | | 0.000 | 00:33 |
| Control Test | | 0.078 | 00:33 |
| Air Blank | | 0.000 | 00:34 |
| Diagnostics Check | | OK | 00:34 |

Data Download Date/Time: 08/19/2019 16:30

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANTWELL,J,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: CANTWELL,J
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: GUSTIN

DATE/TIME OF TEST: 08/15/2019 00:54:23

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:47 |
| Air Blank | | 0.000 | 00:48 |
| Control Test | | 0.078 | 00:48 |
| Air Blank | | 0.000 | 00:49 |
| Subject Sample | # 1 | 0.246 | 00:49 |
| Breath Volume | | 2.691 | 00:49 |
| Air Blank | | 0.000 | 00:50 |
| Air Blank | | 0.000 | 00:52 |
| Subject Sample | # 2 | 0.233 | 00:52 |
| Breath Volume | | 2.175 | 00:52 |
| Air Blank | | 0.000 | 00:53 |
| Control Test | | 0.078 | 00:53 |
| Air Blank | | 0.000 | 00:54 |
| Diagnostics Check | | OK | 00:54 |

Data Download Date/Time: 08/19/2019 16:30

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,SHAUN,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MARTIN,SHAUN
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:02

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 08/15/2019 20:29:18

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 20:22 |
| Air Blank | | 0.000 | 20:23 |
| Control Test | | 0.079 | 20:23 |
| Air Blank | | 0.000 | 20:24 |
| Subject Sample | # 1 | 0.155 | 20:24 |
| Breath Volume | | 3.113 | 20:24 |
| Air Blank | | 0.000 | 20:25 |
| Air Blank | | 0.000 | 20:27 |
| Subject Sample | # 2 | 0.156 | 20:27 |
| Breath Volume | | 2.062 | 20:27 |
| Air Blank | | 0.000 | 20:28 |
| Control Test | | 0.079 | 20:28 |
| Air Blank | | 0.000 | 20:29 |
| Diagnostics Check | | OK | 20:29 |

Data Download Date/Time: 08/19/2019 16:30

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DEP,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MILLER,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: DALLAS

DATE/TIME OF TEST: 08/15/2019 22:48:25

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 22:36 | |
| Air Blank | | 0.000 | 22:37 | |
| Control Test | | 0.078 | 22:37 | |
| Air Blank | | 0.000 | 22:38 | |
| Subject Sample | # 1 | 0.185 | 22:41 | Volume Not Met |
| Breath Volume | | 0.335 | 22:41 | |
| Air Blank | | 0.000 | 22:41 | |
| Air Blank | | 0.000 | 22:43 | |
| Subject Sample | # 2 | 0.199 | 22:46 | Volume Not Met |
| Breath Volume | | 0.324 | 22:46 | |
| Air Blank | | 0.000 | 22:47 | |
| Control Test | | 0.077 | 22:47 | |
| Air Blank | | 0.000 | 22:48 | |
| Diagnostics Check | | OK | 22:48 | |

Data Download Date/Time: 08/19/2019 16:30

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUEHLER,DEP,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BUTLER,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: LEAVITT

DATE/TIME OF TEST: 08/18/2019 02:08:28

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:02 |
| Air Blank | | 0.000 | 02:02 |
| Control Test | | 0.080 | 02:02 |
| Air Blank | | 0.000 | 02:03 |
| Subject Sample | # 1 | 0.185 | 02:03 |
| Breath Volume | | 2.433 | 02:03 |
| Air Blank | | 0.000 | 02:04 |
| Air Blank | | 0.000 | 02:06 |
| Subject Sample | # 2 | 0.189 | 02:07 |
| Breath Volume | | 1.605 | 02:07 |
| Air Blank | | 0.000 | 02:07 |
| Control Test | | 0.080 | 02:07 |
| Air Blank | | 0.000 | 02:08 |
| Diagnostics Check | | OK | 02:08 |

Data Download Date/Time: 08/19/2019 16:30

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUEHLER,DEP,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BOWERS,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: HOMOKY

DATE/TIME OF TEST: 08/18/2019 02:18:52

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:11 |
| Air Blank | | 0.000 | 02:12 |
| Control Test | | 0.079 | 02:12 |
| Air Blank | | 0.000 | 02:13 |
| Subject Sample | # 1 | 0.153 | 02:14 |
| Breath Volume | | 1.769 | 02:14 |
| Air Blank | | 0.000 | 02:14 |
| Air Blank | | 0.000 | 02:16 |
| Subject Sample | # 2 | 0.150 | 02:17 |
| Breath Volume | | 1.484 | 02:17 |
| Air Blank | | 0.000 | 02:17 |
| Control Test | | 0.076 | 02:18 |
| Air Blank | | 0.000 | 02:18 |
| Diagnostics Check | | OK | 02:18 |

Data Download Date/Time: 08/19/2019 16:30

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,NICKKI,N
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: LOWE,B
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: ELLIOTT

DATE/TIME OF TEST: 08/18/2019 05:34:24

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 05:26 |
| Air Blank | | 0.000 | 05:27 |
| Control Test | | 0.079 | 05:27 |
| Air Blank | | 0.000 | 05:28 |
| Subject Sample | # 1 | 0.190 | 05:29 |
| Breath Volume | | 1.382 | 05:29 |
| Air Blank | | 0.000 | 05:29 |
| Air Blank | | 0.000 | 05:31 |
| Subject Sample | # 2 | 0.200 | 05:32 |
| Breath Volume | | 1.511 | 05:32 |
| Air Blank | | 0.000 | 05:33 |
| Control Test | | 0.079 | 05:33 |
| Air Blank | | 0.000 | 05:34 |
| Diagnostics Check | | OK | 05:34 |

Data Download Date/Time: 08/19/2019 16:30

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUEHLER,DEP,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BUEHLER,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: IBARRA

DATE/TIME OF TEST: 08/22/2019 02:18:13

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:11 |
| Air Blank | | 0.000 | 02:12 |
| Control Test | | 0.079 | 02:12 |
| Air Blank | | 0.000 | 02:13 |
| Subject Sample | # 1 | 0.207 | 02:13 |
| Breath Volume | | 2.089 | 02:13 |
| Air Blank | | 0.000 | 02:14 |
| Air Blank | | 0.000 | 02:16 |
| Subject Sample | # 2 | 0.212 | 02:16 |
| Breath Volume | | 1.621 | 02:16 |
| Air Blank | | 0.000 | 02:17 |
| Control Test | | 0.079 | 02:17 |
| Air Blank | | 0.000 | 02:18 |
| Diagnostics Check | | OK | 02:18 |

Data Download Date/Time: 09/25/2019 07:39

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NICK,A,
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: NICK,A
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 08/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 15:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:18

SUBJECT'S LAST NAME: BYNUM

DATE/TIME OF TEST: 08/22/2019 18:41:55

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 18:38 | |
| Air Blank | | 0.000 | 18:39 | |
| Control Test | | 0.080 | 18:39 | |
| Air Blank | | 0.000 | 18:40 | |
| Subject Sample | # 1 | 0.000 | 18:40 | |
| Breath Volume | | 2.320 | 18:40 | |
| Air Blank | | | 18:40 | Improper Sample |
| Air Blank | | 0.000 | 18:41 | |
| Control Test | | 0.081 | 18:41 | |
| Air Blank | | 0.000 | 18:41 | |
| Diagnostics Check | | OK | 18:41 | |

Data Download Date/Time: 09/25/2019 07:39

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NICK,A,
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: NICK,A
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 08/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 15:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:18

SUBJECT'S LAST NAME: BYNUM

DATE/TIME OF TEST: 08/22/2019 18:50:02

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 18:44 |
| Air Blank | | 0.000 | 18:44 |
| Control Test | | 0.080 | 18:44 |
| Air Blank | | 0.000 | 18:45 |
| Subject Sample | # 1 | 0.000 | 18:45 |
| Breath Volume | | 2.718 | 18:45 |
| Air Blank | | 0.000 | 18:46 |
| Air Blank | | 0.000 | 18:48 |
| Subject Sample | # 2 | 0.000 | 18:48 |
| Breath Volume | | 1.335 | 18:48 |
| Air Blank | | 0.000 | 18:49 |
| Control Test | | 0.079 | 18:49 |
| Air Blank | | 0.000 | 18:49 |
| Diagnostics Check | | OK | 18:50 |

Data Download Date/Time: 09/25/2019 07:39

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDEZ,DEPUTY,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEPUTY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: BEST

DATE/TIME OF TEST: 08/23/2019 22:11:05

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 22:05 |
| Air Blank | | 0.000 | 22:05 |
| Control Test | | 0.080 | 22:05 |
| Air Blank | | 0.000 | 22:06 |
| Subject Sample | # 1 | 0.000 | 22:06 |
| Breath Volume | | 4.601 | 22:06 |
| Air Blank | | 0.000 | 22:07 |
| Air Blank | | 0.000 | 22:09 |
| Subject Sample | # 2 | 0.000 | 22:09 |
| Breath Volume | | 5.121 | 22:09 |
| Air Blank | | 0.000 | 22:10 |
| Control Test | | 0.079 | 22:10 |
| Air Blank | | 0.000 | 22:11 |
| Diagnostics Check | | OK | 22:11 |

Data Download Date/Time: 09/25/2019 07:39

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANTWELL,J,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: CANTWELL,TPR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 08/24/2019 22:07:48

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 22:01 |
| Air Blank | | 0.000 | 22:01 |
| Control Test | | 0.080 | 22:01 |
| Air Blank | | 0.000 | 22:02 |
| Subject Sample | # 1 | 0.000 | 22:03 |
| Breath Volume | | 4.410 | 22:03 |
| Air Blank | | 0.000 | 22:03 |
| Air Blank | | 0.000 | 22:05 |
| Subject Sample | # 2 | 0.000 | 22:06 |
| Breath Volume | | 4.203 | 22:06 |
| Air Blank | | 0.000 | 22:06 |
| Control Test | | 0.079 | 22:07 |
| Air Blank | | 0.000 | 22:07 |
| Diagnostics Check | | OK | 22:07 |

Data Download Date/Time: 09/25/2019 07:39

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DEP,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MILLER,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: TRAUTMAN

DATE/TIME OF TEST: 08/25/2019 01:27:52

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:21 |
| Air Blank | | 0.000 | 01:22 |
| Control Test | | 0.079 | 01:22 |
| Air Blank | | 0.000 | 01:22 |
| Subject Sample | # 1 | 0.151 | 01:23 |
| Breath Volume | | 2.457 | 01:23 |
| Air Blank | | 0.000 | 01:23 |
| Air Blank | | 0.000 | 01:25 |
| Subject Sample | # 2 | 0.142 | 01:26 |
| Breath Volume | | 2.398 | 01:26 |
| Air Blank | | 0.000 | 01:26 |
| Control Test | | 0.079 | 01:27 |
| Air Blank | | 0.000 | 01:27 |
| Diagnostics Check | | OK | 01:27 |

Data Download Date/Time: 09/25/2019 07:39

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: ABNER

DATE/TIME OF TEST: 08/25/2019 01:56:19

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:50 |
| Air Blank | | 0.000 | 01:50 |
| Control Test | | 0.080 | 01:51 |
| Air Blank | | 0.000 | 01:51 |
| Subject Sample | # 1 | 0.144 | 01:52 |
| Breath Volume | | 2.078 | 01:52 |
| Air Blank | | 0.000 | 01:52 |
| Air Blank | | 0.000 | 01:54 |
| Subject Sample | # 2 | 0.148 | 01:54 |
| Breath Volume | | 1.308 | 01:54 |
| Air Blank | | 0.000 | 01:55 |
| Control Test | | 0.080 | 01:55 |
| Air Blank | | 0.000 | 01:56 |
| Diagnostics Check | | OK | 01:56 |

Data Download Date/Time: 09/25/2019 07:39

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUMANN,DEP,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BALES,OFC
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 08/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: ZARVES

DATE/TIME OF TEST: 08/25/2019 02:08:26

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:02 |
| Air Blank | | 0.000 | 02:02 |
| Control Test | | 0.078 | 02:02 |
| Air Blank | | 0.000 | 02:03 |
| Subject Sample | # 1 | 0.209 | 02:03 |
| Breath Volume | | 2.550 | 02:03 |
| Air Blank | | 0.000 | 02:04 |
| Air Blank | | 0.000 | 02:06 |
| Subject Sample | # 2 | 0.202 | 02:06 |
| Breath Volume | | 2.292 | 02:06 |
| Air Blank | | 0.000 | 02:07 |
| Control Test | | 0.079 | 02:07 |
| Air Blank | | 0.000 | 02:08 |
| Diagnostics Check | | OK | 02:08 |

Data Download Date/Time: 09/25/2019 07:39

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NICK,A,
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: NICK,A.
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 08/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 08/25/2019 03:29:47

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:23 |
| Air Blank | | 0.000 | 03:24 |
| Control Test | | 0.079 | 03:24 |
| Air Blank | | 0.000 | 03:25 |
| Subject Sample | # 1 | 0.115 | 03:25 |
| Breath Volume | | 3.472 | 03:25 |
| Air Blank | | 0.000 | 03:26 |
| Air Blank | | 0.000 | 03:27 |
| Subject Sample | # 2 | 0.116 | 03:28 |
| Breath Volume | | 2.929 | 03:28 |
| Air Blank | | 0.000 | 03:28 |
| Control Test | | 0.079 | 03:29 |
| Air Blank | | 0.000 | 03:29 |
| Diagnostics Check | | OK | 03:29 |

Data Download Date/Time: 09/25/2019 07:39

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DEP,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: RODRIGUEZ-AGUILON

DATE/TIME OF TEST: 08/25/2019 04:33:16

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:26 |
| Air Blank | | 0.000 | 04:27 |
| Control Test | | 0.080 | 04:27 |
| Air Blank | | 0.000 | 04:27 |
| Subject Sample | # 1 | 0.161 | 04:29 |
| Breath Volume | | 2.570 | 04:29 |
| Air Blank | | 0.000 | 04:29 |
| Air Blank | | 0.000 | 04:31 |
| Subject Sample | # 2 | 0.153 | 04:31 |
| Breath Volume | | 2.136 | 04:31 |
| Air Blank | | 0.000 | 04:32 |
| Control Test | | 0.079 | 04:32 |
| Air Blank | | 0.000 | 04:33 |
| Diagnostics Check | | OK | 04:33 |

Data Download Date/Time: 09/25/2019 07:39

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,DANIEL,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: CAMPBELL,OFC
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 08/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:08

SUBJECT'S LAST NAME: HANKISON

DATE/TIME OF TEST: 08/26/2019 18:11:04

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 18:04 |
| Air Blank | | 0.000 | 18:05 |
| Control Test | | 0.081 | 18:05 |
| Air Blank | | 0.000 | 18:06 |
| Subject Sample | # 1 | 0.035 | 18:06 |
| Breath Volume | | 1.968 | 18:06 |
| Air Blank | | 0.000 | 18:07 |
| Air Blank | | 0.000 | 18:09 |
| Subject Sample | # 2 | 0.033 | 18:09 |
| Breath Volume | | 2.570 | 18:09 |
| Air Blank | | 0.000 | 18:10 |
| Control Test | | 0.079 | 18:10 |
| Air Blank | | 0.000 | 18:10 |
| Diagnostics Check | | OK | 18:11 |

Data Download Date/Time: 09/25/2019 07:39

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLYNN,MICHAEL,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: UCHIDA,Y
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 08/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 08/30/2019 02:25:26

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:19 |
| Air Blank | | 0.000 | 02:19 |
| Control Test | | 0.079 | 02:20 |
| Air Blank | | 0.000 | 02:20 |
| Subject Sample | # 1 | 0.168 | 02:21 |
| Breath Volume | | 2.273 | 02:21 |
| Air Blank | | 0.000 | 02:21 |
| Air Blank | | 0.000 | 02:23 |
| Subject Sample | # 2 | 0.158 | 02:23 |
| Breath Volume | | 2.914 | 02:23 |
| Air Blank | | 0.000 | 02:24 |
| Control Test | | 0.079 | 02:24 |
| Air Blank | | 0.000 | 02:25 |
| Diagnostics Check | | OK | 02:25 |

Data Download Date/Time: 09/25/2019 07:39

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUEHLER,DEP,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BUTLER,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: HOWE

DATE/TIME OF TEST: 08/31/2019 02:53:14

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:46 |
| Air Blank | | 0.000 | 02:47 |
| Control Test | | 0.080 | 02:47 |
| Air Blank | | 0.000 | 02:48 |
| Subject Sample | # 1 | 0.165 | 02:48 |
| Breath Volume | | 2.929 | 02:48 |
| Air Blank | | 0.000 | 02:49 |
| Air Blank | | 0.000 | 02:51 |
| Subject Sample | # 2 | 0.169 | 02:51 |
| Breath Volume | | 3.023 | 02:51 |
| Air Blank | | 0.000 | 02:52 |
| Control Test | | 0.079 | 02:52 |
| Air Blank | | 0.000 | 02:53 |
| Diagnostics Check | | OK | 02:53 |

Data Download Date/Time: 09/25/2019 07:39

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,DANIEL,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BOWERS,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: STRINGER

DATE/TIME OF TEST: 08/31/2019 04:23:33

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:17 |
| Air Blank | | 0.000 | 04:17 |
| Control Test | | 0.080 | 04:18 |
| Air Blank | | 0.000 | 04:18 |
| Subject Sample | # 1 | 0.116 | 04:19 |
| Breath Volume | | 2.554 | 04:19 |
| Air Blank | | 0.000 | 04:19 |
| Air Blank | | 0.000 | 04:21 |
| Subject Sample | # 2 | 0.116 | 04:22 |
| Breath Volume | | 2.703 | 04:22 |
| Air Blank | | 0.000 | 04:22 |
| Control Test | | 0.080 | 04:22 |
| Air Blank | | 0.000 | 04:23 |
| Diagnostics Check | | OK | 04:23 |

Data Download Date/Time: 09/25/2019 07:39

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,DANIEL,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BOWERS,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: STRINGER

DATE/TIME OF TEST: 08/31/2019 05:01:33

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:55 |
| Air Blank | | 0.000 | 04:56 |
| Control Test | | 0.080 | 04:56 |
| Air Blank | | 0.000 | 04:56 |
| Subject Sample | # 1 | 0.100 | 04:57 |
| Breath Volume | | 2.601 | 04:57 |
| Air Blank | | 0.000 | 04:57 |
| Air Blank | | 0.000 | 04:59 |
| Subject Sample | # 2 | 0.101 | 05:00 |
| Breath Volume | | 2.753 | 05:00 |
| Air Blank | | 0.000 | 05:00 |
| Control Test | | 0.080 | 05:00 |
| Air Blank | | 0.000 | 05:01 |
| Diagnostics Check | | OK | 05:01 |

Data Download Date/Time: 09/25/2019 07:39

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,JIMMY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: ADAMS,JIMMY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 08:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:30

SUBJECT'S LAST NAME: MERLO

DATE/TIME OF TEST: 08/31/2019 09:56:52

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 09:50 |
| Air Blank | | 0.000 | 09:51 |
| Control Test | | 0.080 | 09:51 |
| Air Blank | | 0.000 | 09:52 |
| Subject Sample | # 1 | 0.190 | 09:52 |
| Breath Volume | | 3.566 | 09:52 |
| Air Blank | | 0.000 | 09:53 |
| Air Blank | | 0.000 | 09:54 |
| Subject Sample | # 2 | 0.186 | 09:55 |
| Breath Volume | | 2.671 | 09:55 |
| Air Blank | | 0.000 | 09:55 |
| Control Test | | 0.078 | 09:56 |
| Air Blank | | 0.000 | 09:56 |
| Diagnostics Check | | OK | 09:56 |

Data Download Date/Time: 09/25/2019 07:39

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,STEVEN,P
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BUTLER,C
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: DILORENZO

DATE/TIME OF TEST: 08/31/2019 23:03:03

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 22:55 |
| Air Blank | | 0.000 | 22:56 |
| Control Test | | 0.079 | 22:56 |
| Air Blank | | 0.000 | 22:57 |
| Subject Sample | # 1 | 0.087 | 22:58 |
| Breath Volume | | 2.261 | 22:58 |
| Air Blank | | 0.000 | 22:59 |
| Air Blank | | 0.000 | 23:01 |
| Subject Sample | # 2 | 0.078 | 23:01 |
| Breath Volume | | 1.886 | 23:01 |
| Air Blank | | 0.000 | 23:02 |
| Control Test | | 0.078 | 23:02 |
| Air Blank | | 0.000 | 23:02 |
| Diagnostics Check | | OK | 23:03 |

Data Download Date/Time: 09/25/2019 07:39

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK,WILLIAM,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HITCHCOCK,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 539031
DRY GAS EXPIRY DATE: 08/25/2019
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: CRESPO-DELGADO

DATE/TIME OF TEST: 08/02/2019 03:09:49

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:03 |
| Air Blank | | 0.000 | 03:04 |
| Control Test | | 0.079 | 03:04 |
| Air Blank | | 0.000 | 03:05 |
| Subject Sample | # 1 | 0.141 | 03:05 |
| Breath Volume | | 1.843 | 03:05 |
| Air Blank | | 0.000 | 03:06 |
| Air Blank | | 0.000 | 03:07 |
| Subject Sample | # 2 | 0.139 | 03:08 |
| Breath Volume | | 2.316 | 03:08 |
| Air Blank | | 0.000 | 03:08 |
| Control Test | | 0.079 | 03:09 |
| Air Blank | | 0.000 | 03:09 |
| Diagnostics Check | | OK | 03:09 |

Data Download Date/Time: 08/23/2019 21:55

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK',WILLIAM,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HITCHCOCK,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 539031
DRY GAS EXPIRY DATE: 08/25/2019
VIOLATION CODE: DUI
ARREST TIME: 04:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: NICOL

DATE/TIME OF TEST: 08/06/2019 05:13:12

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 05:06 |
| Air Blank | | 0.000 | 05:07 |
| Control Test | | 0.079 | 05:07 |
| Air Blank | | 0.000 | 05:07 |
| Subject Sample | # 1 | 0.000 | 05:09 |
| Breath Volume | | 1.238 | 05:09 |
| Air Blank | | 0.000 | 05:09 |
| Air Blank | | 0.000 | 05:11 |
| Subject Sample | # 2 | 0.000 | 05:11 |
| Breath Volume | | 1.308 | 05:11 |
| Air Blank | | 0.000 | 05:12 |
| Control Test | | 0.079 | 05:12 |
| Air Blank | | 0.000 | 05:13 |
| Diagnostics Check | | OK | 05:13 |

Data Download Date/Time: 08/23/2019 21:55

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK,WILLIAM,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HITCHCOCK,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 539031
DRY GAS EXPIRY DATE: 08/25/2019
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: ASHLEY

DATE/TIME OF TEST: 08/10/2019 00:40:18

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:34 |
| Air Blank | | 0.000 | 00:34 |
| Control Test | | 0.079 | 00:34 |
| Air Blank | | 0.000 | 00:35 |
| Subject Sample | # 1 | 0.102 | 00:36 |
| Breath Volume | | 2.378 | 00:36 |
| Air Blank | | 0.000 | 00:36 |
| Air Blank | | 0.000 | 00:38 |
| Subject Sample | # 2 | 0.102 | 00:38 |
| Breath Volume | | 2.511 | 00:38 |
| Air Blank | | 0.000 | 00:39 |
| Control Test | | 0.079 | 00:39 |
| Air Blank | | 0.000 | 00:40 |
| Diagnostics Check | | OK | 00:40 |

Data Download Date/Time: 08/23/2019 21:55

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANLEY,JEFFREY,A
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: STANLEY,JEFFREY
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 539031
DRY GAS EXPIRY DATE: 08/25/2019
VIOLATION CODE: DUI
ARREST TIME: 15:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:40

SUBJECT'S LAST NAME: VINCENT

DATE/TIME OF TEST: 08/10/2019 16:17:58

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 16:11 |
| Air Blank | | 0.000 | 16:12 |
| Control Test | | 0.079 | 16:12 |
| Air Blank | | 0.000 | 16:12 |
| Subject Sample | # 1 | 0.000 | 16:13 |
| Breath Volume | | 1.949 | 16:13 |
| Air Blank | | 0.000 | 16:14 |
| Air Blank | | 0.000 | 16:16 |
| Subject Sample | # 2 | 0.000 | 16:16 |
| Breath Volume | | 2.953 | 16:16 |
| Air Blank | | 0.000 | 16:17 |
| Control Test | | 0.080 | 16:17 |
| Air Blank | | 0.000 | 16:17 |
| Diagnostics Check | | OK | 16:17 |

Data Download Date/Time: 08/23/2019 21:55

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK,WILLIAM,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HITCHCOCK,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: SANDOVAL MARTINEZ

DATE/TIME OF TEST: 08/23/2019 04:41:50

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:35 |
| Air Blank | | 0.000 | 04:36 |
| Control Test | | 0.080 | 04:36 |
| Air Blank | | 0.000 | 04:37 |
| Subject Sample | # 1 | 0.019 | 04:37 |
| Breath Volume | | 1.390 | 04:37 |
| Air Blank | | 0.000 | 04:38 |
| Air Blank | | 0.000 | 04:40 |
| Subject Sample | # 2 | 0.018 | 04:40 |
| Breath Volume | | 2.218 | 04:40 |
| Air Blank | | 0.000 | 04:40 |
| Control Test | | 0.081 | 04:41 |
| Air Blank | | 0.000 | 04:41 |
| Diagnostics Check | | OK | 04:41 |

Data Download Date/Time: 08/23/2019 21:55

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK,WILLIAM,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HITCHCOCK,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: BROWN JR

DATE/TIME OF TEST: 08/25/2019 01:30:38

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:23 |
| Air Blank | | 0.000 | 01:24 |
| Control Test | | 0.080 | 01:24 |
| Air Blank | | 0.000 | 01:25 |
| Subject Sample | # 1 | 0.000 | 01:25 |
| Breath Volume | | 1.167 | 01:25 |
| Air Blank | | 0.000 | 01:25 |
| Air Blank | | 0.000 | 01:27 |
| Subject Sample | # 2 | 0.000 | 01:29 |
| Breath Volume | | 1.164 | 01:29 |
| Air Blank | | 0.000 | 01:29 |
| Control Test | | 0.082 | 01:30 |
| Air Blank | | 0.000 | 01:30 |
| Diagnostics Check | | OK | 01:30 |

Data Download Date/Time: 09/30/2019 22:04

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,CHANNING,L
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: TAYLOR,CHANNING
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 05:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:20

SUBJECT'S LAST NAME: STRIDIRON

DATE/TIME OF TEST: 08/29/2019 06:59:36

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 06:51 |
| Air Blank | | 0.000 | 06:52 |
| Control Test | | 0.080 | 06:52 |
| Air Blank | | 0.000 | 06:52 |
| Subject Sample | # 1 | 0.135 | 06:55 |
| Breath Volume | | 1.332 | 06:55 |
| Air Blank | | 0.000 | 06:56 |
| Air Blank | | 0.000 | 06:57 |
| Subject Sample | # 2 | 0.137 | 06:58 |
| Breath Volume | | 1.359 | 06:58 |
| Air Blank | | 0.000 | 06:58 |
| Control Test | | 0.080 | 06:59 |
| Air Blank | | 0.000 | 06:59 |
| Diagnostics Check | | OK | 06:59 |

Data Download Date/Time: 09/30/2019 22:04

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVARD,TERESA,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SOLTIS,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 07/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: RIVARD

DATE/TIME OF TEST: 08/05/2019 23:06:03

| | | | |
|---|---|---|---|
| Diagnostics Check | OK | 23:05 | |
| Air Blank | | 23:05 | Improper Sample |
| Air Blank | | 23:06 | Improper Sample |

Data Download Date/Time: 08/18/2019 00:37


COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZAMORA,DEPUTY,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: KNAPP,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 07/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: OGLINE

DATE/TIME OF TEST: 08/06/2019 03:36:31

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:30 |
| Air Blank | | 0.000 | 03:30 |
| Control Test | | 0.080 | 03:31 |
| Air Blank | | 0.000 | 03:31 |
| Subject Sample | # 1 | 0.180 | 03:32 |
| Breath Volume | | 2.152 | 03:32 |
| Air Blank | | 0.000 | 03:32 |
| Air Blank | | 0.000 | 03:34 |
| Subject Sample | # 2 | 0.176 | 03:35 |
| Breath Volume | | 2.597 | 03:35 |
| Air Blank | | 0.000 | 03:35 |
| Control Test | | 0.079 | 03:35 |
| Air Blank | | 0.000 | 03:36 |
| Diagnostics Check | | OK | 03:36 |

Data Download Date/Time: 08/18/2019 00:37

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZAMORA,DEPUTY,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: FRASER,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 07/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: PHILIPS

DATE/TIME OF TEST: 08/10/2019 23:37:29

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:31 |
| Air Blank | | 0.000 | 23:31 |
| Control Test | | 0.079 | 23:31 |
| Air Blank | | 0.000 | 23:32 |
| Subject Sample | # 1 | 0.296 | 23:32 |
| Breath Volume | | 2.230 | 23:32 |
| Air Blank | | 0.000 | 23:33 |
| Air Blank | | 0.000 | 23:35 |
| Subject Sample | # 2 | 0.288 | 23:35 |
| Breath Volume | | 2.785 | 23:35 |
| Air Blank | | 0.000 | 23:36 |
| Control Test | | 0.079 | 23:36 |
| Air Blank | | 0.000 | 23:37 |
| Diagnostics Check | | OK | 23:37 |

Data Download Date/Time: 08/18/2019 00:37

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RONGA,DAVID,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: HOLLINGSWORTH,CODY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 07/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: HAILER

DATE/TIME OF TEST: 08/16/2019 02:12:02

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:04 |
| Air Blank | | 0.000 | 02:05 |
| Control Test | | 0.079 | 02:05 |
| Air Blank | | 0.000 | 02:06 |
| Subject Sample | # 1 | 0.213 | 02:07 |
| Breath Volume | | 1.226 | 02:07 |
| Air Blank | | 0.000 | 02:07 |
| Air Blank | | 0.000 | 02:09 |
| Subject Sample | # 2 | 0.195 | 02:10 |
| Breath Volume | | 1.585 | 02:10 |
| Air Blank | | 0.000 | 02:11 |
| Control Test | | 0.080 | 02:11 |
| Air Blank | | 0.000 | 02:11 |
| Diagnostics Check | | OK | 02:12 |

Data Download Date/Time: 08/18/2019 00:37

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZAMORA,DEPUTY,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: LAROSA,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 07/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: DOROBIALA

DATE/TIME OF TEST: 08/16/2019 02:32:53

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:26 |
| Air Blank | | 0.000 | 02:27 |
| Control Test | | 0.079 | 02:27 |
| Air Blank | | 0.000 | 02:27 |
| Subject Sample | # 1 | 0.239 | 02:28 |
| Breath Volume | | 2.937 | 02:28 |
| Air Blank | | 0.000 | 02:28 |
| Air Blank | | 0.000 | 02:30 |
| Subject Sample | # 2 | 0.245 | 02:31 |
| Breath Volume | | 3.199 | 02:31 |
| Air Blank | | 0.000 | 02:31 |
| Control Test | | 0.079 | 02:32 |
| Air Blank | | 0.000 | 02:32 |
| Diagnostics Check | | OK | 02:32 |

Data Download Date/Time: 08/18/2019 00:37

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

INSTRUMENT REGISTERED TO: NAS MAYPORT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHAHEEN,KEITH,B
BREATH TEST OPERATOR AGENCY: Mayport Naval Base
ARRESTING OFFICER: NA,NA
ARRESTING OFFICER AGENCY: Mayport Naval Base
LAST AGENCY INSPECTION: 07/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: ADMIN
ARREST TIME: 10:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:15

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 08/04/2019 11:01:22

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 10:55 |
| Air Blank | | 0.000 | 10:55 |
| Control Test | | 0.079 | 10:55 |
| Air Blank | | 0.000 | 10:56 |
| Subject Sample | # 1 | 0.000 | 10:57 |
| Breath Volume | | 4.539 | 10:57 |
| Air Blank | | 0.000 | 10:57 |
| Air Blank | | 0.000 | 10:59 |
| Subject Sample | # 2 | 0.000 | 11:00 |
| Breath Volume | | 4.488 | 11:00 |
| Air Blank | | 0.000 | 11:00 |
| Control Test | | 0.078 | 11:00 |
| Air Blank | | 0.000 | 11:01 |
| Diagnostics Check | | OK | 11:01 |

Data Download Date/Time: 08/16/2019 14:33

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

INSTRUMENT REGISTERED TO: NAS MAYPORT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OCASIO,JOSE,A
BREATH TEST OPERATOR AGENCY: Mayport Naval Base
ARRESTING OFFICER: LOPEZ,RAYMOND
ARRESTING OFFICER AGENCY: Mayport Naval Base
LAST AGENCY INSPECTION: 07/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 08/11/2019 04:05:40

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 03:51 | |
| Air Blank | | 0.000 | 03:51 | |
| Control Test | | 0.080 | 03:52 | |
| Air Blank | | 0.000 | 03:52 | |
| Subject Sample | # 1 | 0.054 | 03:53 | |
| Breath Volume | | 2.843 | 03:53 | |
| Air Blank | | 0.000 | 03:53 | |
| Air Blank | | 0.000 | 03:55 | |
| Subject Sample | # 2 | | 03:58 | No Sample Provided |
| Breath Volume | | 0.000 | 03:58 | |
| Air Blank | | 0.000 | 03:59 | |
| Air Blank | | 0.000 | 04:01 | |
| Subject Sample | # 3 | | 04:04 | No Sample Provided |
| Breath Volume | | 0.000 | 04:04 | |
| Air Blank | | 0.000 | 04:04 | |
| Control Test | | 0.080 | 04:05 | |
| Air Blank | | 0.000 | 04:05 | |
| Diagnostics Check | | OK | 04:05 | |

Data Download Date/Time: 08/16/2019 14:33

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

INSTRUMENT REGISTERED TO: NAS MAYPORT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OCASIO,JOSE,A
BREATH TEST OPERATOR AGENCY: Mayport Naval Base
ARRESTING OFFICER: LOPEZ,RAYMOND
ARRESTING OFFICER AGENCY: Mayport Naval Base
LAST AGENCY INSPECTION: 07/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 08/11/2019 04:14:50

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:08 |
| Air Blank | | 0.000 | 04:09 |
| Control Test | | 0.080 | 04:09 |
| Air Blank | | 0.000 | 04:10 |
| Subject Sample | # 1 | 0.050 | 04:10 |
| Breath Volume | | 2.726 | 04:10 |
| Air Blank | | 0.000 | 04:11 |
| Air Blank | | 0.000 | 04:12 |
| Subject Sample | # 2 | 0.050 | 04:13 |
| Breath Volume | | 2.925 | 04:13 |
| Air Blank | | 0.000 | 04:13 |
| Control Test | | 0.080 | 04:14 |
| Air Blank | | 0.000 | 04:14 |
| Diagnostics Check | | OK | 04:14 |

Data Download Date/Time: 08/16/2019 14:33

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JOHNS,MARK
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 07/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 05:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:36

SUBJECT'S LAST NAME: WINDHAM

DATE/TIME OF TEST: 08/04/2019 06:06:26

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 06:00 |
| Air Blank | | 0.000 | 06:00 |
| Control Test | | 0.082 | 06:01 |
| Air Blank | | 0.000 | 06:01 |
| Subject Sample | # 1 | 0.187 | 06:02 |
| Breath Volume | | 2.710 | 06:02 |
| Air Blank | | 0.000 | 06:02 |
| Air Blank | | 0.000 | 06:04 |
| Subject Sample | # 2 | 0.183 | 06:04 |
| Breath Volume | | 1.945 | 06:04 |
| Air Blank | | 0.000 | 06:05 |
| Control Test | | 0.081 | 06:05 |
| Air Blank | | 0.000 | 06:06 |
| Diagnostics Check | | OK | 06:06 |

Data Download Date/Time: 08/07/2019 14:54

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SHEIMAN,RICHARD
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 07/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: VELEZ HERNANDEZ

DATE/TIME OF TEST: 08/05/2019 00:34:31

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:28 |
| Air Blank | | 0.000 | 00:28 |
| Control Test | | 0.081 | 00:28 |
| Air Blank | | 0.000 | 00:29 |
| Subject Sample | # 1 | 0.021 | 00:30 |
| Breath Volume | | 3.167 | 00:30 |
| Air Blank | | 0.000 | 00:30 |
| Air Blank | | 0.000 | 00:32 |
| Subject Sample | # 2 | 0.019 | 00:33 |
| Breath Volume | | 2.105 | 00:33 |
| Air Blank | | 0.000 | 00:33 |
| Control Test | | 0.080 | 00:33 |
| Air Blank | | 0.000 | 00:34 |
| Diagnostics Check | | OK | 00:34 |

Data Download Date/Time: 08/07/2019 14:54

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CORREA,ASHLEY
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 07/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: DOORLEY

DATE/TIME OF TEST: 08/13/2019 03:15:05

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:06 |
| Air Blank | | 0.000 | 03:07 |
| Control Test | | 0.081 | 03:07 |
| Air Blank | | 0.000 | 03:08 |
| Subject Sample | # 1 | 0.211 | 03:09 |
| Breath Volume | | 3.414 | 03:09 |
| Air Blank | | 0.000 | 03:09 |
| Air Blank | | 0.000 | 03:11 |
| Subject Sample | # 2 | 0.201 | 03:13 |
| Breath Volume | | 2.500 | 03:13 |
| Air Blank | | 0.000 | 03:14 |
| Control Test | | 0.080 | 03:14 |
| Air Blank | | 0.000 | 03:14 |
| Diagnostics Check | | OK | 03:15 |

Data Download Date/Time: 08/15/2019 10:20

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELVILLE,KELLY,M
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JOHNSON,MICHAEL
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 08/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 14:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:12

SUBJECT'S LAST NAME: BOTERO

DATE/TIME OF TEST: 08/15/2019 15:41:59

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 15:36 |
| Air Blank | | 0.000 | 15:36 |
| Control Test | | 0.080 | 15:36 |
| Air Blank | | 0.000 | 15:37 |
| Subject Sample | # 1 | 0.344 | 15:37 |
| Breath Volume | | 1.351 | 15:37 |
| Air Blank | | 0.000 | 15:38 |
| Air Blank | | 0.000 | 15:40 |
| Subject Sample | # 2 | 0.339 | 15:40 |
| Breath Volume | | 1.152 | 15:40 |
| Air Blank | | 0.000 | 15:41 |
| Control Test | | 0.080 | 15:41 |
| Air Blank | | 0.000 | 15:41 |
| Diagnostics Check | | OK | 15:41 |

Data Download Date/Time: 08/22/2019 13:39

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RIVERA,VICTOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: MCDOWELL

DATE/TIME OF TEST: 08/16/2019 00:11:17

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:04 |
| Air Blank | | 0.000 | 00:05 |
| Control Test | | 0.080 | 00:05 |
| Air Blank | | 0.000 | 00:06 |
| Subject Sample | # 1 | 0.142 | 00:06 |
| Breath Volume | | 3.394 | 00:06 |
| Air Blank | | 0.000 | 00:07 |
| Air Blank | | 0.000 | 00:09 |
| Subject Sample | # 2 | 0.138 | 00:09 |
| Breath Volume | | 3.000 | 00:09 |
| Air Blank | | 0.000 | 00:10 |
| Control Test | | 0.080 | 00:10 |
| Air Blank | | 0.000 | 00:11 |
| Diagnostics Check | | OK | 00:11 |

Data Download Date/Time: 08/22/2019 13:39

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JUNCO,JONATHAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 08/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: TOLENTINO

DATE/TIME OF TEST: 08/16/2019 02:09:19

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:01 |
| Air Blank | | 0.000 | 02:01 |
| Control Test | | 0.080 | 02:02 |
| Air Blank | | 0.000 | 02:02 |
| Subject Sample | # 1 | 0.230 | 02:04 |
| Breath Volume | | 3.242 | 02:04 |
| Air Blank | | 0.000 | 02:04 |
| Air Blank | | 0.000 | 02:06 |
| Subject Sample | # 2 | 0.231 | 02:07 |
| Breath Volume | | 3.839 | 02:07 |
| Air Blank | | 0.000 | 02:08 |
| Control Test | | 0.080 | 02:08 |
| Air Blank | | 0.000 | 02:09 |
| Diagnostics Check | | OK | 02:09 |

Data Download Date/Time: 08/22/2019 13:39

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HARWOOD,ARMANDO
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 08/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: ARBOLEDA

DATE/TIME OF TEST: 08/16/2019 04:18:50

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:12 |
| Air Blank | | 0.000 | 04:12 |
| Control Test | | 0.080 | 04:13 |
| Air Blank | | 0.000 | 04:13 |
| Subject Sample | # 1 | 0.109 | 04:14 |
| Breath Volume | | 2.320 | 04:14 |
| Air Blank | | 0.000 | 04:15 |
| Air Blank | | 0.000 | 04:16 |
| Subject Sample | # 2 | 0.105 | 04:17 |
| Breath Volume | | 1.531 | 04:17 |
| Air Blank | | 0.000 | 04:17 |
| Control Test | | 0.080 | 04:18 |
| Air Blank | | 0.000 | 04:18 |
| Diagnostics Check | | OK | 04:18 |

Data Download Date/Time: 08/22/2019 13:40

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MORALES,JOSE
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 08/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: GRAHAM

DATE/TIME OF TEST: 08/17/2019 01:21:32

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 01:18 | |
| Air Blank | | 0.000 | 01:19 | |
| Control Test | | 0.080 | 01:19 | |
| Air Blank | | 0.000 | 01:19 | |
| Subject Sample | # 1 | | 01:20 | Subject Test Refused |
| Breath Volume | | 0.000 | 01:20 | |
| Air Blank | | 0.000 | 01:20 | |
| Control Test | | 0.081 | 01:20 | |
| Air Blank | | 0.000 | 01:21 | |
| Diagnostics Check | | OK | 01:21 | |

Data Download Date/Time: 08/22/2019 13:40

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KERN,ETHAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 08/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:49

SUBJECT'S LAST NAME: FATTORE

DATE/TIME OF TEST: 08/17/2019 21:18:20

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 21:11 |
| Air Blank | | 0.000 | 21:12 |
| Control Test | | 0.080 | 21:12 |
| Air Blank | | 0.000 | 21:13 |
| Subject Sample | # 1 | 0.174 | 21:13 |
| Breath Volume | | 1.343 | 21:13 |
| Air Blank | | 0.000 | 21:14 |
| Air Blank | | 0.000 | 21:15 |
| Subject Sample | # 2 | 0.191 | 21:16 |
| Breath Volume | | 2.328 | 21:16 |
| Air Blank | | 0.000 | 21:17 |
| Control Test | | 0.080 | 21:17 |
| Air Blank | | 0.000 | 21:18 |
| Diagnostics Check | | OK | 21:18 |

Data Download Date/Time: 08/22/2019 13:40

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AMARAL,WILKER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: NAVARRO ESTALA

DATE/TIME OF TEST: 08/17/2019 22:39:03

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 22:32 |
| Air Blank | | 0.000 | 22:32 |
| Control Test | | 0.081 | 22:32 |
| Air Blank | | 0.000 | 22:33 |
| Subject Sample | # 1 | 0.194 | 22:34 |
| Breath Volume | | 4.410 | 22:34 |
| Air Blank | | 0.000 | 22:35 |
| Air Blank | | 0.000 | 22:36 |
| Subject Sample | # 2 | 0.194 | 22:37 |
| Breath Volume | | 3.875 | 22:37 |
| Air Blank | | 0.000 | 22:38 |
| Control Test | | 0.080 | 22:38 |
| Air Blank | | 0.000 | 22:38 |
| Diagnostics Check | | OK | 22:39 |

Data Download Date/Time: 08/22/2019 13:40

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DOUGHERTY,JOSEPH
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 08/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: ZAHRADNIK

DATE/TIME OF TEST: 08/18/2019 00:37:14

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 00:32 | |
| Air Blank | | 0.000 | 00:33 | |
| Control Test | | 0.081 | 00:33 | |
| Air Blank | | 0.000 | 00:33 | |
| Subject Sample | # 1 | | 00:35 | Subject Test Refused |
| Breath Volume | | 0.000 | 00:35 | |
| Air Blank | | 0.000 | 00:36 | |
| Control Test | | 0.081 | 00:36 | |
| Air Blank | | 0.000 | 00:37 | |
| Diagnostics Check | | OK | 00:37 | |

Data Download Date/Time: 08/22/2019 13:40

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JOHNS,MARK
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 08/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 08/18/2019 02:11:06

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 01:58 | |
| Air Blank | | 0.000 | 01:58 | |
| Control Test | | 0.081 | 01:58 | |
| Air Blank | | 0.000 | 01:59 | |
| Subject Sample | # 1 | 0.179 | 02:02 | Volume Not Met |
| Breath Volume | | 0.000 | 02:02 | |
| Air Blank | | 0.000 | 02:03 | |
| Air Blank | | 0.000 | 02:05 | |
| Subject Sample | # 2 | 0.252 | 02:05 | No .020 agreement |
| Breath Volume | | 2.035 | 02:05 | |
| Air Blank | | 0.000 | 02:06 | |
| Air Blank | | 0.000 | 02:08 | |
| Subject Sample | # 3 | 0.228 | 02:09 | No .020 agreement |
| Breath Volume | | 1.925 | 02:09 | |
| Air Blank | | 0.000 | 02:10 | |
| Control Test | | 0.080 | 02:10 | |
| Air Blank | | 0.000 | 02:11 | |
| Diagnostics Check | | OK | 02:11 | |

Data Download Date/Time: 08/22/2019 13:40

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JOHNS,MARK
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 08/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 08/18/2019 02:27:06

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:18 |
| Air Blank | | 0.000 | 02:19 |
| Control Test | | 0.080 | 02:19 |
| Air Blank | | 0.000 | 02:20 |
| Subject Sample | # 1 | 0.239 | 02:22 |
| Breath Volume | | 2.707 | 02:22 |
| Air Blank | | 0.000 | 02:23 |
| Air Blank | | 0.000 | 02:25 |
| Subject Sample | # 2 | 0.229 | 02:25 |
| Breath Volume | | 2.476 | 02:25 |
| Air Blank | | 0.000 | 02:26 |
| Control Test | | 0.077 | 02:26 |
| Air Blank | | 0.000 | 02:27 |
| Diagnostics Check | | OK | 02:27 |

Data Download Date/Time: 08/22/2019 13:40

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MARINI,TAYLOR
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 08/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: ROSENBERG

DATE/TIME OF TEST: 08/18/2019 03:46:12

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 03:34 | |
| Air Blank | | 0.000 | 03:35 | |
| Control Test | | 0.080 | 03:35 | |
| Air Blank | | 0.000 | 03:36 | |
| Subject Sample | # 1 | | 03:39 | No Sample Provided |
| Breath Volume | | 0.000 | 03:39 | |
| Air Blank | | 0.000 | 03:39 | |
| Air Blank | | 0.000 | 03:41 | |
| Subject Sample | # 2 | | 03:44 | No Sample Provided |
| Breath Volume | | 0.000 | 03:44 | |
| Air Blank | | 0.000 | 03:45 | |
| Control Test | | 0.081 | 03:45 | |
| Air Blank | | 0.000 | 03:46 | |
| Diagnostics Check | | OK | 03:46 | |

Data Download Date/Time: 08/22/2019 13:40

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MONTGOMERY,JAMES
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 08/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: FISHER

DATE/TIME OF TEST: 08/18/2019 04:58:18

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:51 |
| Air Blank | | 0.000 | 04:51 |
| Control Test | | 0.081 | 04:51 |
| Air Blank | | 0.000 | 04:52 |
| Subject Sample | # 1 | 0.158 | 04:52 |
| Breath Volume | | 2.988 | 04:52 |
| Air Blank | | 0.000 | 04:53 |
| Air Blank | | 0.000 | 04:55 |
| Subject Sample | # 2 | 0.162 | 04:56 |
| Breath Volume | | 2.457 | 04:56 |
| Air Blank | | 0.000 | 04:57 |
| Control Test | | 0.081 | 04:57 |
| Air Blank | | 0.000 | 04:58 |
| Diagnostics Check | | OK | 04:58 |

Data Download Date/Time: 08/22/2019 13:40

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PRICE,FORREST
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 08/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: SULIVERAS

DATE/TIME OF TEST: 08/18/2019 22:16:52

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 22:10 |
| Air Blank | | 0.000 | 22:11 |
| Control Test | | 0.081 | 22:11 |
| Air Blank | | 0.000 | 22:12 |
| Subject Sample | # 1 | 0.197 | 22:12 |
| Breath Volume | | 3.578 | 22:12 |
| Air Blank | | 0.000 | 22:13 |
| Air Blank | | 0.000 | 22:15 |
| Subject Sample | # 2 | 0.179 | 22:15 |
| Breath Volume | | 1.156 | 22:15 |
| Air Blank | | 0.000 | 22:15 |
| Control Test | | 0.080 | 22:16 |
| Air Blank | | 0.000 | 22:16 |
| Diagnostics Check | | OK | 22:16 |

Data Download Date/Time: 08/22/2019 13:40

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RIVERA,HERIBERTO
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 08/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 08/19/2019 23:32:19

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:26 |
| Air Blank | | 0.000 | 23:26 |
| Control Test | | 0.081 | 23:27 |
| Air Blank | | 0.000 | 23:27 |
| Subject Sample | # 1 | 0.000 | 23:28 |
| Breath Volume | | 1.570 | 23:28 |
| Air Blank | | 0.000 | 23:28 |
| Air Blank | | 0.000 | 23:30 |
| Subject Sample | # 2 | 0.000 | 23:30 |
| Breath Volume | | 2.273 | 23:30 |
| Air Blank | | 0.000 | 23:31 |
| Control Test | | 0.080 | 23:31 |
| Air Blank | | 0.000 | 23:32 |
| Diagnostics Check | | OK | 23:32 |

Data Download Date/Time: 08/22/2019 13:40

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RODRIGUEZ,JORGE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 08/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: LEBOWITZ

DATE/TIME OF TEST: 08/20/2019 21:30:50

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 21:23 |
| Air Blank | | 0.000 | 21:23 |
| Control Test | | 0.081 | 21:23 |
| Air Blank | | 0.000 | 21:24 |
| Subject Sample | # 1 | 0.141 | 21:26 |
| Breath Volume | | 1.867 | 21:26 |
| Air Blank | | 0.000 | 21:26 |
| Air Blank | | 0.000 | 21:28 |
| Subject Sample | # 2 | 0.136 | 21:29 |
| Breath Volume | | 2.371 | 21:29 |
| Air Blank | | 0.000 | 21:29 |
| Control Test | | 0.080 | 21:30 |
| Air Blank | | 0.000 | 21:30 |
| Diagnostics Check | | OK | 21:30 |

Data Download Date/Time: 08/22/2019 13:40

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CORTES,ADAM
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 08/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 08/22/2019 04:30:08

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:22 |
| Air Blank | | 0.000 | 04:23 |
| Control Test | | 0.080 | 04:23 |
| Air Blank | | 0.000 | 04:23 |
| Subject Sample | # 1 | 0.208 | 04:25 |
| Breath Volume | | 2.437 | 04:25 |
| Air Blank | | 0.000 | 04:25 |
| Air Blank | | 0.000 | 04:27 |
| Subject Sample | # 2 | 0.212 | 04:28 |
| Breath Volume | | 2.558 | 04:28 |
| Air Blank | | 0.000 | 04:29 |
| Control Test | | 0.079 | 04:29 |
| Air Blank | | 0.000 | 04:30 |
| Diagnostics Check | | OK | 04:30 |

Data Download Date/Time: 08/22/2019 13:40

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MONTGOMERY,JAMES
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 08/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: DOBIS

DATE/TIME OF TEST: 08/23/2019 23:24:27

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:16 |
| Air Blank | | 0.000 | 23:17 |
| Control Test | | 0.081 | 23:17 |
| Air Blank | | 0.000 | 23:18 |
| Subject Sample | # 1 | 0.201 | 23:19 |
| Breath Volume | | 3.750 | 23:19 |
| Air Blank | | 0.000 | 23:20 |
| Air Blank | | 0.000 | 23:22 |
| Subject Sample | # 2 | 0.196 | 23:22 |
| Breath Volume | | 3.339 | 23:22 |
| Air Blank | | 0.000 | 23:23 |
| Control Test | | 0.081 | 23:23 |
| Air Blank | | 0.000 | 23:24 |
| Diagnostics Check | | OK | 23:24 |

Data Download Date/Time: 08/29/2019 16:38

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AUSTIN,ADAM
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 08/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: SHEPPARD

DATE/TIME OF TEST: 08/24/2019 22:43:00

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 22:34 |
| Air Blank | | 0.000 | 22:35 |
| Control Test | | 0.081 | 22:35 |
| Air Blank | | 0.000 | 22:35 |
| Subject Sample | # 1 | 0.204 | 22:37 |
| Breath Volume | | 1.695 | 22:37 |
| Air Blank | | 0.000 | 22:38 |
| Air Blank | | 0.000 | 22:40 |
| Subject Sample | # 2 | 0.215 | 22:41 |
| Breath Volume | | 2.308 | 22:41 |
| Air Blank | | 0.000 | 22:42 |
| Control Test | | 0.079 | 22:42 |
| Air Blank | | 0.000 | 22:42 |
| Diagnostics Check | | OK | 22:43 |

Data Download Date/Time: 08/29/2019 16:38

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RODRIGUEZ,STEPHANIE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 08/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: WEBB

DATE/TIME OF TEST: 08/24/2019 23:27:58

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:21 |
| Air Blank | | 0.000 | 23:22 |
| Control Test | | 0.080 | 23:22 |
| Air Blank | | 0.000 | 23:23 |
| Subject Sample | # 1 | 0.244 | 23:23 |
| Breath Volume | | 4.320 | 23:23 |
| Air Blank | | 0.000 | 23:24 |
| Air Blank | | 0.000 | 23:26 |
| Subject Sample | # 2 | 0.237 | 23:26 |
| Breath Volume | | 3.035 | 23:26 |
| Air Blank | | 0.000 | 23:27 |
| Control Test | | 0.080 | 23:27 |
| Air Blank | | 0.000 | 23:27 |
| Diagnostics Check | | OK | 23:27 |

Data Download Date/Time: 08/29/2019 16:39

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CSIZMADIA,ANDRAS
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 08/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: HAWLEY

DATE/TIME OF TEST: 08/25/2019 02:19:56

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:13 |
| Air Blank | | 0.000 | 02:13 |
| Control Test | | 0.080 | 02:14 |
| Air Blank | | 0.000 | 02:14 |
| Subject Sample | # 1 | 0.197 | 02:15 |
| Breath Volume | | 3.332 | 02:15 |
| Air Blank | | 0.000 | 02:16 |
| Air Blank | | 0.000 | 02:17 |
| Subject Sample | # 2 | 0.193 | 02:18 |
| Breath Volume | | 1.207 | 02:18 |
| Air Blank | | 0.000 | 02:19 |
| Control Test | | 0.079 | 02:19 |
| Air Blank | | 0.000 | 02:19 |
| Diagnostics Check | | OK | 02:19 |

Data Download Date/Time: 08/29/2019 16:39

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BIRDSONG,KEVIN
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 08/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: HENSON

DATE/TIME OF TEST: 08/25/2019 03:27:40

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:21 |
| Air Blank | | 0.000 | 03:21 |
| Control Test | | 0.081 | 03:21 |
| Air Blank | | 0.000 | 03:22 |
| Subject Sample | # 1 | 0.167 | 03:23 |
| Breath Volume | | 3.488 | 03:23 |
| Air Blank | | 0.000 | 03:23 |
| Air Blank | | 0.000 | 03:25 |
| Subject Sample | # 2 | 0.163 | 03:26 |
| Breath Volume | | 3.460 | 03:26 |
| Air Blank | | 0.000 | 03:26 |
| Control Test | | 0.079 | 03:27 |
| Air Blank | | 0.000 | 03:27 |
| Diagnostics Check | | OK | 03:27 |

Data Download Date/Time: 08/29/2019 16:39

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: FAIRBANKS,ANTHONY
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 08/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: MICKLEY

DATE/TIME OF TEST: 08/25/2019 03:47:33

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:41 |
| Air Blank | | 0.000 | 03:41 |
| Control Test | | 0.079 | 03:42 |
| Air Blank | | 0.000 | 03:42 |
| Subject Sample | # 1 | 0.169 | 03:43 |
| Breath Volume | | 4.324 | 03:43 |
| Air Blank | | 0.000 | 03:43 |
| Air Blank | | 0.000 | 03:45 |
| Subject Sample | # 2 | 0.172 | 03:46 |
| Breath Volume | | 4.394 | 03:46 |
| Air Blank | | 0.000 | 03:46 |
| Control Test | | 0.080 | 03:46 |
| Air Blank | | 0.000 | 03:47 |
| Diagnostics Check | | OK | 03:47 |

Data Download Date/Time: 08/29/2019 16:39

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HARRIS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 08/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:58

SUBJECT'S LAST NAME: QUINTANA

DATE/TIME OF TEST: 08/25/2019 05:25:11

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 05:21 | |
| Air Blank | | 0.000 | 05:22 | |
| Control Test | | 0.081 | 05:22 | |
| Air Blank | | 0.000 | 05:22 | |
| Subject Sample | # 1 | | 05:23 | Subject Test Refused |
| Breath Volume | | 0.000 | 05:23 | |
| Air Blank | | 0.000 | 05:24 | |
| Control Test | | 0.080 | 05:24 | |
| Air Blank | | 0.000 | 05:25 | |
| Diagnostics Check | | OK | 05:25 | |

Data Download Date/Time: 08/29/2019 16:39

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LANE,KEATON
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 08/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:49

SUBJECT'S LAST NAME: RODRIGUEZREYES

DATE/TIME OF TEST: 08/25/2019 20:18:35

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 20:12 |
| Air Blank | | 0.000 | 20:12 |
| Control Test | | 0.081 | 20:12 |
| Air Blank | | 0.000 | 20:13 |
| Subject Sample | # 1 | 0.263 | 20:13 |
| Breath Volume | | 1.226 | 20:13 |
| Air Blank | | 0.000 | 20:14 |
| Air Blank | | 0.000 | 20:16 |
| Subject Sample | # 2 | 0.244 | 20:17 |
| Breath Volume | | 1.289 | 20:17 |
| Air Blank | | 0.000 | 20:17 |
| Control Test | | 0.079 | 20:17 |
| Air Blank | | 0.000 | 20:18 |
| Diagnostics Check | | OK | 20:18 |

Data Download Date/Time: 08/29/2019 16:39

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BARBER,ALEX
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 08/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: BOZEK

DATE/TIME OF TEST: 08/27/2019 03:58:46

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:52 |
| Air Blank | | 0.000 | 03:52 |
| Control Test | | 0.081 | 03:53 |
| Air Blank | | 0.000 | 03:53 |
| Subject Sample | # 1 | 0.194 | 03:54 |
| Breath Volume | | 4.343 | 03:54 |
| Air Blank | | 0.000 | 03:54 |
| Air Blank | | 0.000 | 03:56 |
| Subject Sample | # 2 | 0.195 | 03:57 |
| Breath Volume | | 4.039 | 03:57 |
| Air Blank | | 0.000 | 03:57 |
| Control Test | | 0.080 | 03:58 |
| Air Blank | | 0.000 | 03:58 |
| Diagnostics Check | | OK | 03:58 |

Data Download Date/Time: 08/29/2019 16:39

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WILLIAMS,JUSTIN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 08/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 08/28/2019 01:19:55

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:08 |
| Air Blank | | 0.000 | 01:08 |
| Control Test | | 0.081 | 01:08 |
| Air Blank | | 0.000 | 01:09 |
| Subject Sample | # 1 | 0.215 | 01:10 |
| Breath Volume | | 2.417 | 01:10 |
| Air Blank | | 0.000 | 01:11 |
| Air Blank | | 0.000 | 01:12 |
| Subject Sample | # 2 | 0.190 | 01:14 |
| Breath Volume | | 1.140 | 01:14 |
| Air Blank | | 0.000 | 01:15 |
| Air Blank | | 0.000 | 01:17 |
| Subject Sample | # 3 | 0.189 | 01:18 |
| Breath Volume | | 1.441 | 01:18 |
| Air Blank | | 0.000 | 01:19 |
| Control Test | | 0.079 | 01:19 |
| Air Blank | | 0.000 | 01:19 |
| Diagnostics Check | | OK | 01:19 |

Data Download Date/Time: 08/29/2019 16:39

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LOPEZ,OSCAR
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 08/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: ATKINS

DATE/TIME OF TEST: 08/28/2019 02:46:29

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:39 |
| Air Blank | | 0.000 | 02:39 |
| Control Test | | 0.080 | 02:40 |
| Air Blank | | 0.000 | 02:40 |
| Subject Sample | # 1 | 0.144 | 02:41 |
| Breath Volume | | 3.136 | 02:41 |
| Air Blank | | 0.000 | 02:42 |
| Air Blank | | 0.000 | 02:44 |
| Subject Sample | # 2 | 0.138 | 02:44 |
| Breath Volume | | 2.195 | 02:44 |
| Air Blank | | 0.000 | 02:45 |
| Control Test | | 0.080 | 02:45 |
| Air Blank | | 0.000 | 02:46 |
| Diagnostics Check | | OK | 02:46 |

Data Download Date/Time: 08/29/2019 16:39

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RODRIGUEZ,JORGE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 08/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: VARGAS

DATE/TIME OF TEST: 08/29/2019 02:54:07

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:44 |
| Air Blank | | 0.000 | 02:45 |
| Control Test | | 0.080 | 02:45 |
| Air Blank | | 0.000 | 02:46 |
| Subject Sample | # 1 | 0.234 | 02:46 |
| Breath Volume | | 2.398 | 02:46 |
| Air Blank | | 0.000 | 02:47 |
| Air Blank | | 0.000 | 02:49 |
| Subject Sample | # 2 | 0.212 | 02:49 |
| Breath Volume | | 1.515 | 02:49 |
| Air Blank | | 0.000 | 02:50 |
| Air Blank | | 0.000 | 02:52 |
| Subject Sample | # 3 | 0.240 | 02:52 |
| Breath Volume | | 2.078 | 02:52 |
| Air Blank | | 0.000 | 02:53 |
| Control Test | | 0.079 | 02:53 |
| Air Blank | | 0.000 | 02:54 |
| Diagnostics Check | | OK | 02:54 |

Data Download Date/Time: 08/29/2019 16:39

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,ARMAND,G
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JANESEK,BRANDON
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 08/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 07:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:13

SUBJECT'S LAST NAME: UNDA

DATE/TIME OF TEST: 08/29/2019 08:44:59

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 08:38 |
| Air Blank | | 0.000 | 08:38 |
| Control Test | | 0.082 | 08:39 |
| Air Blank | | 0.000 | 08:39 |
| Subject Sample | # 1 | 0.000 | 08:40 |
| Breath Volume | | 4.558 | 08:40 |
| Air Blank | | 0.000 | 08:40 |
| Air Blank | | 0.000 | 08:42 |
| Subject Sample | # 2 | 0.000 | 08:43 |
| Breath Volume | | 3.761 | 08:43 |
| Air Blank | | 0.000 | 08:44 |
| Control Test | | 0.080 | 08:44 |
| Air Blank | | 0.000 | 08:44 |
| Diagnostics Check | | OK | 08:44 |

Data Download Date/Time: 08/29/2019 16:39

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: STARNES,JACOB
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 08/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: PAYNE

DATE/TIME OF TEST: 08/30/2019 01:04:20

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:56 |
| Air Blank | | 0.000 | 00:56 |
| Control Test | | 0.081 | 00:57 |
| Air Blank | | 0.000 | 00:57 |
| Subject Sample | # 1 | 0.255 | 00:58 |
| Breath Volume | | 1.324 | 00:58 |
| Air Blank | | 0.000 | 00:59 |
| Air Blank | | 0.000 | 01:01 |
| Subject Sample | # 2 | 0.259 | 01:02 |
| Breath Volume | | 1.425 | 01:02 |
| Air Blank | | 0.000 | 01:03 |
| Control Test | | 0.079 | 01:03 |
| Air Blank | | 0.000 | 01:04 |
| Diagnostics Check | | OK | 01:04 |

Data Download Date/Time: 09/04/2019 14:59

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JORGE,JUNIOR
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 08/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 08/30/2019 01:52:15

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:45 |
| Air Blank | | 0.000 | 01:46 |
| Control Test | | 0.081 | 01:46 |
| Air Blank | | 0.000 | 01:46 |
| Subject Sample | # 1 | 0.000 | 01:47 |
| Breath Volume | | 2.312 | 01:47 |
| Air Blank | | 0.000 | 01:48 |
| Air Blank | | 0.000 | 01:50 |
| Subject Sample | # 2 | 0.000 | 01:50 |
| Breath Volume | | 4.175 | 01:50 |
| Air Blank | | 0.000 | 01:51 |
| Control Test | | 0.081 | 01:51 |
| Air Blank | | 0.000 | 01:52 |
| Diagnostics Check | | OK | 01:52 |

Data Download Date/Time: 09/04/2019 14:59

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KIRK,BRIAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 08/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: BARBAL

DATE/TIME OF TEST: 08/30/2019 03:16:36

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:10 |
| Air Blank | | 0.000 | 03:10 |
| Control Test | | 0.081 | 03:10 |
| Air Blank | | 0.000 | 03:11 |
| Subject Sample | # 1 | 0.000 | 03:12 |
| Breath Volume | | 2.285 | 03:12 |
| Air Blank | | 0.000 | 03:12 |
| Air Blank | | 0.000 | 03:14 |
| Subject Sample | # 2 | 0.000 | 03:15 |
| Breath Volume | | 1.660 | 03:15 |
| Air Blank | | 0.000 | 03:15 |
| Control Test | | 0.082 | 03:16 |
| Air Blank | | 0.000 | 03:16 |
| Diagnostics Check | | OK | 03:16 |

Data Download Date/Time: 09/04/2019 14:59

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TUERO,ADRIAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 07:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:25

SUBJECT'S LAST NAME: LITTLE

DATE/TIME OF TEST: 08/30/2019 09:54:18

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 09:48 |
| Air Blank | | 0.000 | 09:48 |
| Control Test | | 0.081 | 09:49 |
| Air Blank | | 0.000 | 09:49 |
| Subject Sample | # 1 | 0.000 | 09:50 |
| Breath Volume | | 1.687 | 09:50 |
| Air Blank | | 0.000 | 09:50 |
| Air Blank | | 0.000 | 09:52 |
| Subject Sample | # 2 | 0.000 | 09:52 |
| Breath Volume | | 1.246 | 09:52 |
| Air Blank | | 0.000 | 09:53 |
| Control Test | | 0.081 | 09:53 |
| Air Blank | | 0.000 | 09:54 |
| Diagnostics Check | | OK | 09:54 |

Data Download Date/Time: 09/04/2019 14:59

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GUEVARA,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 12:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:25

SUBJECT'S LAST NAME: HOLSEN

DATE/TIME OF TEST: 08/30/2019 14:01:58

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 13:55 |
| Air Blank | | 0.000 | 13:55 |
| Control Test | | 0.081 | 13:56 |
| Air Blank | | 0.000 | 13:56 |
| Subject Sample | # 1 | 0.000 | 13:57 |
| Breath Volume | | 1.812 | 13:57 |
| Air Blank | | 0.000 | 13:57 |
| Air Blank | | 0.000 | 14:00 |
| Subject Sample | # 2 | 0.000 | 14:00 |
| Breath Volume | | 1.277 | 14:00 |
| Air Blank | | 0.000 | 14:01 |
| Control Test | | 0.081 | 14:01 |
| Air Blank | | 0.000 | 14:01 |
| Diagnostics Check | | OK | 14:01 |

Data Download Date/Time: 09/04/2019 14:59

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BRIGGS,DERIENNE
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 08/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: MCKIBBEN

DATE/TIME OF TEST: 08/30/2019 22:56:42

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 22:49 |
| Air Blank | | 0.000 | 22:50 |
| Control Test | | 0.082 | 22:50 |
| Air Blank | | 0.000 | 22:51 |
| Subject Sample | # 1 | 0.160 | 22:51 |
| Breath Volume | | 2.410 | 22:51 |
| Air Blank | | 0.000 | 22:52 |
| Air Blank | | 0.000 | 22:54 |
| Subject Sample | # 2 | 0.151 | 22:55 |
| Breath Volume | | 1.312 | 22:55 |
| Air Blank | | 0.000 | 22:55 |
| Control Test | | 0.080 | 22:56 |
| Air Blank | | 0.000 | 22:56 |
| Diagnostics Check | | OK | 22:56 |

Data Download Date/Time: 09/04/2019 14:59

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RAMIREZ,MATTHEW
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 08/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: KOMANECK

DATE/TIME OF TEST: 08/30/2019 23:45:40

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:38 |
| Air Blank | | 0.000 | 23:38 |
| Control Test | | 0.081 | 23:39 |
| Air Blank | | 0.000 | 23:39 |
| Subject Sample | # 1 | 0.229 | 23:41 |
| Breath Volume | | 1.417 | 23:41 |
| Air Blank | | 0.000 | 23:41 |
| Air Blank | | 0.000 | 23:43 |
| Subject Sample | # 2 | 0.231 | 23:44 |
| Breath Volume | | 1.449 | 23:44 |
| Air Blank | | 0.000 | 23:44 |
| Control Test | | 0.079 | 23:45 |
| Air Blank | | 0.000 | 23:45 |
| Diagnostics Check | | OK | 23:45 |

Data Download Date/Time: 09/04/2019 14:59

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DOMINGUEZ,JAYMES
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 08/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: HERNANDEZDEJESUS

DATE/TIME OF TEST: 08/31/2019 00:40:07

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:34 |
| Air Blank | | 0.000 | 00:34 |
| Control Test | | 0.082 | 00:34 |
| Air Blank | | 0.000 | 00:35 |
| Subject Sample | # 1 | 0.000 | 00:35 |
| Breath Volume | | 3.148 | 00:35 |
| Air Blank | | 0.000 | 00:36 |
| Air Blank | | 0.000 | 00:38 |
| Subject Sample | # 2 | 0.000 | 00:38 |
| Breath Volume | | 2.722 | 00:38 |
| Air Blank | | 0.000 | 00:39 |
| Control Test | | 0.081 | 00:39 |
| Air Blank | | 0.000 | 00:40 |
| Diagnostics Check | | OK | 00:40 |

Data Download Date/Time: 09/04/2019 14:59

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPLAN,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: DACEY,MICHAEL
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: STENGER JR

DATE/TIME OF TEST: 08/20/2019 00:11:53

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:05 |
| Air Blank | | 0.000 | 00:05 |
| Control Test | | 0.075 | 00:06 |
| Air Blank | | 0.000 | 00:06 |
| Subject Sample | # 1 | 0.172 | 00:07 |
| Breath Volume | | 2.867 | 00:07 |
| Air Blank | | 0.000 | 00:08 |
| Air Blank | | 0.000 | 00:09 |
| Subject Sample | # 2 | 0.173 | 00:10 |
| Breath Volume | | 2.648 | 00:10 |
| Air Blank | | 0.000 | 00:10 |
| Control Test | | 0.076 | 00:11 |
| Air Blank | | 0.000 | 00:11 |
| Diagnostics Check | | OK | 00:11 |

Data Download Date/Time: 08/25/2019 21:49

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001646

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUBNIS,GUY,W
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: GARCIA,MICHAEL
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 07/29/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 08/02/2019 01:44:44

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:35 |
| Air Blank | | 0.000 | 01:36 |
| Control Test | | 0.081 | 01:36 |
| Air Blank | | 0.000 | 01:36 |
| Subject Sample | # 1 | 0.091 | 01:37 |
| Breath Volume | | 1.242 | 01:37 |
| Air Blank | | 0.000 | 01:38 |
| Air Blank | | 0.000 | 01:40 |
| Subject Sample | # 2 | 0.100 | 01:43 |
| Breath Volume | | 1.214 | 01:43 |
| Air Blank | | 0.000 | 01:43 |
| Control Test | | 0.081 | 01:44 |
| Air Blank | | 0.000 | 01:44 |
| Diagnostics Check | | OK | 01:44 |

Data Download Date/Time: 09/04/2019 10:50

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001646

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUBNIS,GUY,W
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: MARTIN,TIFFANI
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 07/29/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: OTHER
ARREST TIME: 02:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: LOVE

DATE/TIME OF TEST: 08/02/2019 03:43:49

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 03:29 | |
| Air Blank | | 0.000 | 03:30 | |
| Control Test | | 0.080 | 03:30 | |
| Air Blank | | 0.000 | 03:31 | |
| Subject Sample | # 1 | 0.000 | 03:34 | Volume Not Met |
| Breath Volume | | 0.476 | 03:34 | |
| Air Blank | | 0.000 | 03:34 | |
| Air Blank | | 0.000 | 03:36 | |
| Subject Sample | # 2 | 0.000 | 03:39 | |
| Breath Volume | | 1.722 | 03:39 | |
| Air Blank | | 0.000 | 03:40 | |
| Air Blank | | 0.000 | 03:42 | |
| Subject Sample | # 3 | 0.000 | 03:42 | |
| Breath Volume | | 1.484 | 03:42 | |
| Air Blank | | 0.000 | 03:42 | |
| Control Test | | 0.081 | 03:43 | |
| Air Blank | | 0.000 | 03:43 | |
| Diagnostics Check | | OK | 03:43 | |

Data Download Date/Time: 09/04/2019 10:51

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001646

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOHORQUEZ,RAPHAEL,A
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: GLEASON,JASON
ARRESTING OFFICER AGENCY: Parker PD
LAST AGENCY INSPECTION: 07/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 11:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:26

SUBJECT'S LAST NAME: BIBBS

DATE/TIME OF TEST: 08/06/2019 12:11:02

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 12:03 |
| Air Blank | | 0.000 | 12:04 |
| Control Test | | 0.081 | 12:04 |
| Air Blank | | 0.000 | 12:05 |
| Subject Sample | # 1 | 0.083 | 12:06 |
| Breath Volume | | 3.035 | 12:06 |
| Air Blank | | 0.000 | 12:07 |
| Air Blank | | 0.000 | 12:08 |
| Subject Sample | # 2 | 0.078 | 12:09 |
| Breath Volume | | 3.007 | 12:09 |
| Air Blank | | 0.000 | 12:10 |
| Control Test | | 0.080 | 12:10 |
| Air Blank | | 0.000 | 12:10 |
| Diagnostics Check | | OK | 12:11 |

Data Download Date/Time: 09/04/2019 10:51

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001646

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMONSON,DEREK,W
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: RIVERA,JUAN
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 07/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 08/24/2019 03:19:00

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:12 |
| Air Blank | | 0.000 | 03:12 |
| Control Test | | 0.082 | 03:13 |
| Air Blank | | 0.000 | 03:13 |
| Subject Sample | # 1 | 0.246 | 03:14 |
| Breath Volume | | 1.773 | 03:14 |
| Air Blank | | 0.000 | 03:14 |
| Air Blank | | 0.000 | 03:16 |
| Subject Sample | # 2 | 0.243 | 03:17 |
| Breath Volume | | 2.050 | 03:17 |
| Air Blank | | 0.000 | 03:18 |
| Control Test | | 0.080 | 03:18 |
| Air Blank | | 0.000 | 03:18 |
| Diagnostics Check | | OK | 03:19 |

Data Download Date/Time: 09/04/2019 10:51

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001646

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMONSON,DEREK,W
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: MARTIN,NICOLE
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 07/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: FENNELL

DATE/TIME OF TEST: 08/25/2019 04:25:08

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:16 |
| Air Blank | | 0.000 | 04:16 |
| Control Test | | 0.082 | 04:16 |
| Air Blank | | 0.000 | 04:17 |
| Subject Sample | # 1 | 0.057 | 04:20 |
| Breath Volume | | 1.472 | 04:20 |
| Air Blank | | 0.000 | 04:21 |
| Air Blank | | 0.000 | 04:22 |
| Subject Sample | # 2 | 0.059 | 04:23 |
| Breath Volume | | 1.382 | 04:23 |
| Air Blank | | 0.000 | 04:24 |
| Control Test | | 0.081 | 04:24 |
| Air Blank | | 0.000 | 04:25 |
| Diagnostics Check | | OK | 04:25 |

Data Download Date/Time: 09/04/2019 10:51

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANE,AMANDA,M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: LANE,AMANDA
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 07/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: PATTERSON

DATE/TIME OF TEST: 08/03/2019 02:47:54

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:40 |
| Air Blank | | 0.000 | 02:41 |
| Control Test | | 0.082 | 02:41 |
| Air Blank | | 0.000 | 02:41 |
| Subject Sample | # 1 | 0.120 | 02:43 |
| Breath Volume | | 1.160 | 02:43 |
| Air Blank | | 0.000 | 02:43 |
| Air Blank | | 0.000 | 02:45 |
| Subject Sample | # 2 | 0.116 | 02:46 |
| Breath Volume | | 1.910 | 02:46 |
| Air Blank | | 0.000 | 02:47 |
| Control Test | | 0.081 | 02:47 |
| Air Blank | | 0.000 | 02:47 |
| Diagnostics Check | | OK | 02:47 |

Data Download Date/Time: 08/08/2019 13:02

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRESTON,MIKAYLA,D
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: LITTLE,DAVID
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 07/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 18:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: NELSON-BRINKLEY

DATE/TIME OF TEST: 08/03/2019 19:26:19

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 19:20 |
| Air Blank | | 0.000 | 19:20 |
| Control Test | | 0.082 | 19:21 |
| Air Blank | | 0.000 | 19:21 |
| Subject Sample | # 1 | 0.171 | 19:21 |
| Breath Volume | | 3.175 | 19:21 |
| Air Blank | | 0.000 | 19:22 |
| Air Blank | | 0.000 | 19:24 |
| Subject Sample | # 2 | 0.175 | 19:24 |
| Breath Volume | | 1.765 | 19:24 |
| Air Blank | | 0.000 | 19:25 |
| Control Test | | 0.082 | 19:25 |
| Air Blank | | 0.000 | 19:26 |
| Diagnostics Check | | OK | 19:26 |

Data Download Date/Time: 08/08/2019 13:02

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,TIMOTHY,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: WHITE,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 07/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: RECINOS

DATE/TIME OF TEST: 08/05/2019 22:26:26

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 22:11 | |
| Air Blank | | 0.000 | 22:12 | |
| Control Test | | 0.082 | 22:12 | |
| Air Blank | | 0.000 | 22:13 | |
| Subject Sample | # 1 | | 22:16 | No Sample Provided |
| Breath Volume | | 0.000 | 22:16 | |
| Air Blank | | 0.000 | 22:16 | |
| Air Blank | | 0.000 | 22:18 | |
| Subject Sample | # 2 | 0.156 | 22:19 | |
| Breath Volume | | 1.164 | 22:19 | |
| Air Blank | | 0.000 | 22:19 | |
| Air Blank | | 0.000 | 22:21 | |
| Subject Sample | # 3 | 0.134 | 22:24 | Volume Not Met |
| Breath Volume | | 0.621 | 22:24 | |
| Air Blank | | 0.000 | 22:25 | |
| Control Test | | 0.082 | 22:25 | |
| Air Blank | | 0.000 | 22:26 | |
| Diagnostics Check | | OK | 22:26 | |

Data Download Date/Time: 08/08/2019 13:02

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: APOINTE,A
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 07/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 15:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:45

SUBJECT'S LAST NAME: REINSMITH

DATE/TIME OF TEST: 08/06/2019 16:33:20

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 16:30 | |
| Air Blank | | 0.000 | 16:31 | |
| Control Test | | 0.082 | 16:31 | |
| Air Blank | | 0.000 | 16:31 | |
| Subject Sample | # 1 | | 16:31 | Subject Test Refused |
| Breath Volume | | 0.000 | 16:31 | |
| Air Blank | | 0.000 | 16:32 | |
| Control Test | | 0.082 | 16:32 | |
| Air Blank | | 0.000 | 16:33 | |
| Diagnostics Check | | OK | 16:33 | |

Data Download Date/Time: 08/08/2019 13:02

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001657

INSTRUMENT REGISTERED TO: MICCOSUKEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELENDEZ,DOUGLAS,
BREATH TEST OPERATOR AGENCY: MICCOSUKEE PD
ARRESTING OFFICER: BROWER,RYAN
ARRESTING OFFICER AGENCY: MICCOSUKEE PD
LAST AGENCY INSPECTION: 08/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A2
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: CARDENAS GARCIA

DATE/TIME OF TEST: 08/25/2019 06:37:47

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 06:24 | |
| Air Blank | | 0.000 | 06:24 | |
| Control Test | | 0.075 | 06:25 | |
| Air Blank | | 0.000 | 06:25 | |
| Subject Sample | # 1 | 0.119 | 06:28 | Volume Not Met |
| Breath Volume | | 0.203 | 06:28 | |
| Air Blank | | 0.000 | 06:29 | |
| Air Blank | | 0.000 | 06:31 | |
| Subject Sample | # 2 | 0.158 | 06:32 | |
| Breath Volume | | 2.148 | 06:32 | |
| Air Blank | | 0.000 | 06:33 | |
| Air Blank | | 0.000 | 06:35 | |
| Subject Sample | # 3 | 0.160 | 06:36 | |
| Breath Volume | | 2.250 | 06:36 | |
| Air Blank | | 0.000 | 06:36 | |
| Control Test | | 0.075 | 06:37 | |
| Air Blank | | 0.000 | 06:37 | |
| Diagnostics Check | | OK | 06:37 | |

Data Download Date/Time: 09/19/2019 10:37

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001714

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMILLAN,SHIKINA,N
BREATH TEST OPERATOR AGENCY: U of Fl PD
ARRESTING OFFICER: WILLIAMS,STEPHANIE
ARRESTING OFFICER AGENCY: U of Fl PD
LAST AGENCY INSPECTION: 07/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: FAULK

DATE/TIME OF TEST: 08/10/2019 05:04:57

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 04:52 | |
| Air Blank | | 0.000 | 04:52 | |
| Control Test | | 0.077 | 04:53 | |
| Air Blank | | 0.000 | 04:53 | |
| Subject Sample | # 1 | 0.007 | 04:56 | Volume Not Met |
| Breath Volume | | 0.472 | 04:56 | |
| Air Blank | | 0.000 | 04:57 | |
| Air Blank | | 0.000 | 04:59 | |
| Subject Sample | # 2 | 0.000 | 04:59 | |
| Breath Volume | | 2.148 | 04:59 | |
| Air Blank | | 0.000 | 05:00 | |
| Air Blank | | 0.000 | 05:02 | |
| Subject Sample | # 3 | 0.000 | 05:03 | |
| Breath Volume | | 2.449 | 05:03 | |
| Air Blank | | 0.000 | 05:04 | |
| Control Test | | 0.079 | 05:04 | |
| Air Blank | | 0.000 | 05:04 | |
| Diagnostics Check | | OK | 05:04 | |

Data Download Date/Time: 08/30/2019 15:24

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001714

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMILLAN,SHIKINA,N
BREATH TEST OPERATOR AGENCY: U of Fl PD
ARRESTING OFFICER: WILLIAMS,STEPHANIE
ARRESTING OFFICER AGENCY: U of Fl PD
LAST AGENCY INSPECTION: 07/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: CANADY

DATE/TIME OF TEST: 08/24/2019 04:16:22

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 04:05 | |
| Air Blank | | 0.000 | 04:05 | |
| Control Test | | 0.076 | 04:05 | |
| Air Blank | | 0.000 | 04:06 | |
| Subject Sample | # 1 | | 04:09 | No Sample Provided |
| Breath Volume | | 0.000 | 04:09 | |
| Air Blank | | 0.000 | 04:09 | |
| Air Blank | | 0.000 | 04:11 | |
| Subject Sample | # 2 | | 04:15 | No Sample Provided |
| Breath Volume | | 0.000 | 04:15 | |
| Air Blank | | 0.000 | 04:15 | |
| Control Test | | 0.077 | 04:15 | |
| Air Blank | | 0.000 | 04:16 | |
| Diagnostics Check | | OK | 04:16 | |

Data Download Date/Time: 08/30/2019 15:24

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001714

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMILLAN,SHIKINA,N
BREATH TEST OPERATOR AGENCY: U of Fl PD
ARRESTING OFFICER: WILLIAMS,STEPHANIE
ARRESTING OFFICER AGENCY: U of Fl PD
LAST AGENCY INSPECTION: 07/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: CANDELA

DATE/TIME OF TEST: 08/30/2019 03:32:46

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:25 |
| Air Blank | | 0.000 | 03:25 |
| Control Test | | 0.077 | 03:25 |
| Air Blank | | 0.000 | 03:26 |
| Subject Sample | # 1 | 0.029 | 03:28 |
| Breath Volume | | 2.824 | 03:28 |
| Air Blank | | 0.000 | 03:28 |
| Air Blank | | 0.000 | 03:30 |
| Subject Sample | # 2 | 0.026 | 03:31 |
| Breath Volume | | 2.503 | 03:31 |
| Air Blank | | 0.000 | 03:31 |
| Control Test | | 0.076 | 03:32 |
| Air Blank | | 0.000 | 03:32 |
| Diagnostics Check | | OK | 03:32 |

Data Download Date/Time: 08/30/2019 15:24

**Printed: 21 November 2019**

COBRA ©1997-2014 CMI, Inc.