

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LUKSA,ANDREW,  
BREATH TEST OPERATOR AGENCY: Coral Gables PD  
ARRESTING OFFICER: SARRADET,MICHELLE  
ARRESTING OFFICER AGENCY: Coral Gables PD  
LAST AGENCY INSPECTION: 07/29/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 620051  
DRY GAS EXPIRY DATE: 11/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: ALPENDRE

DATE/TIME OF TEST: 08/03/2019 03:21:55

---

Diagnostics Check	OK	03:19	
Air Blank	0.000	03:19	
Control Test	0.082	03:19	
Air Blank	0.000	03:20	
Subject Sample	# 1	03:20	Subject Test Refused
Breath Volume	0.000	03:20	
Air Blank	0.000	03:21	
Control Test	0.082	03:21	
Air Blank	0.000	03:21	
Diagnostics Check	OK	03:21	

---

Data Download Date/Time: 09/05/2019 10:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BECKER,ALEXA,C  
BREATH TEST OPERATOR AGENCY: Coral Gables PD  
ARRESTING OFFICER: JAMES,ALEXIA  
ARRESTING OFFICER AGENCY: Coral Gables PD  
LAST AGENCY INSPECTION: 07/29/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 620051  
DRY GAS EXPIRY DATE: 11/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:11  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 08/10/2019 03:53:13

---

Diagnostics Check	OK	03:50	
Air Blank	0.000	03:50	
Control Test	0.082	03:51	
Air Blank	0.000	03:51	
Subject Sample	# 1	03:51	Subject Test Refused
Breath Volume	0.000	03:51	
Air Blank	0.000	03:52	
Control Test	0.082	03:52	
Air Blank	0.000	03:53	
Diagnostics Check	OK	03:53	

---

Data Download Date/Time: 09/05/2019 10:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARDIS,JOHN,D  
BREATH TEST OPERATOR AGENCY: Coral Gables PD  
ARRESTING OFFICER: MARDIS,JOHN  
ARRESTING OFFICER AGENCY: Coral Gables PD  
LAST AGENCY INSPECTION: 07/29/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 620051  
DRY GAS EXPIRY DATE: 11/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 08:37  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 08:45

SUBJECT'S LAST NAME: ALFARO TURCIOS

DATE/TIME OF TEST: 08/11/2019 09:34:55

---

Diagnostics Check	OK	09:28
Air Blank	0.000	09:28
Control Test	0.083	09:29
Air Blank	0.000	09:29
Subject Sample # 1	0.206	09:30
Breath Volume	1.402	09:30
Air Blank	0.000	09:31
Air Blank	0.000	09:32
Subject Sample # 2	0.199	09:33
Breath Volume	1.207	09:33
Air Blank	0.000	09:34
Control Test	0.081	09:34
Air Blank	0.000	09:34
Diagnostics Check	OK	09:34

---

Data Download Date/Time: 09/05/2019 10:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PUGA,JORGE,A  
BREATH TEST OPERATOR AGENCY: Coral Gables PD  
ARRESTING OFFICER: GONZALEZ,JUSTIN  
ARRESTING OFFICER AGENCY: Coral Gables PD  
LAST AGENCY INSPECTION: 07/29/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 620051  
DRY GAS EXPIRY DATE: 11/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: HERMIJO

DATE/TIME OF TEST: 08/16/2019 22:30:34

---

Diagnostics Check	OK	22:27	
Air Blank	0.000	22:28	
Control Test	0.082	22:28	
Air Blank	0.000	22:29	
Subject Sample	# 1	22:29	Subject Test Refused
Breath Volume	0.000	22:29	
Air Blank	0.000	22:29	
Control Test	0.082	22:29	
Air Blank	0.000	22:30	
Diagnostics Check	OK	22:30	

---

Data Download Date/Time: 09/05/2019 10:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PUGA,JORGE,A  
BREATH TEST OPERATOR AGENCY: Coral Gables PD  
ARRESTING OFFICER: MAIZO,BRYAN  
ARRESTING OFFICER AGENCY: Coral Gables PD  
LAST AGENCY INSPECTION: 07/29/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 620051  
DRY GAS EXPIRY DATE: 11/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:12  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: MILAN

DATE/TIME OF TEST: 08/22/2019 00:07:28

---

Diagnostics Check	OK	00:04	
Air Blank	0.000	00:05	
Control Test	0.081	00:05	
Air Blank	0.000	00:05	
Subject Sample	# 1	00:06	Subject Test Refused
Breath Volume	0.000	00:06	
Air Blank	0.000	00:06	
Control Test	0.082	00:06	
Air Blank	0.000	00:07	
Diagnostics Check	OK	00:07	

---

Data Download Date/Time: 09/05/2019 10:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BECKER,ALEXA,C  
BREATH TEST OPERATOR AGENCY: Coral Gables PD  
ARRESTING OFFICER: BECKER,ALEXA  
ARRESTING OFFICER AGENCY: Coral Gables PD  
LAST AGENCY INSPECTION: 07/29/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 620051  
DRY GAS EXPIRY DATE: 11/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 12:55

SUBJECT'S LAST NAME: COLMENARES

DATE/TIME OF TEST: 08/22/2019 01:08:34

---

Diagnostics Check	OK	00:57	
Air Blank	0.000	00:57	
Control Test	0.082	00:57	
Air Blank	0.000	00:58	
Subject Sample	# 1	01:01	No Sample Provided
Breath Volume	0.000	01:01	
Air Blank	0.000	01:02	
Air Blank	0.000	01:04	
Subject Sample	# 2	01:07	No Sample Provided
Breath Volume	0.000	01:07	
Air Blank	0.000	01:07	
Control Test	0.081	01:07	
Air Blank	0.000	01:08	
Diagnostics Check	OK	01:08	

---

Data Download Date/Time: 09/05/2019 10:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BECKER,ALEXA,C  
BREATH TEST OPERATOR AGENCY: Coral Gables PD  
ARRESTING OFFICER: BECKER,ALEXA  
ARRESTING OFFICER AGENCY: Coral Gables PD  
LAST AGENCY INSPECTION: 07/29/2019

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 620051  
DRY GAS EXPIRY DATE: 11/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: COLMENARES

DATE/TIME OF TEST: 08/22/2019 01:24:54

---

Diagnostics Check	OK	01:21	
Air Blank	0.000	01:22	
Control Test	0.082	01:22	
Air Blank	0.000	01:22	
Subject Sample	# 1	01:23	Slope Not Met
Breath Volume	1.101	01:23	
Air Blank	0.000	01:23	
Control Test	0.082	01:24	
Air Blank	0.000	01:24	
Diagnostics Check	OK	01:24	

---

Data Download Date/Time: 09/05/2019 10:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BECKER,ALEXA,C  
BREATH TEST OPERATOR AGENCY: Coral Gables PD  
ARRESTING OFFICER: BECKER,ALEXA  
ARRESTING OFFICER AGENCY: Coral Gables PD  
LAST AGENCY INSPECTION: 07/29/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 620051  
DRY GAS EXPIRY DATE: 11/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: COLMENARES

DATE/TIME OF TEST: 08/22/2019 01:32:55

---

Diagnostics Check		OK	01:26
Air Blank		0.000	01:27
Control Test		0.082	01:27
Air Blank		0.000	01:27
Subject Sample	# 1	0.380	01:28
Breath Volume		2.738	01:28
Air Blank		0.000	01:29
Air Blank		0.000	01:30
Subject Sample	# 2	0.376	01:31
Breath Volume		2.257	01:31
Air Blank		0.000	01:32
Control Test		0.080	01:32
Air Blank		0.000	01:32
Diagnostics Check		OK	01:32

---

Data Download Date/Time: 09/05/2019 10:20



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CABELLO,ADRIAN,  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: GOMEZ,JOSEPH  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 07/26/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 640976  
DRY GAS EXPIRY DATE: 11/21/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:56  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: LUNA

DATE/TIME OF TEST: 08/01/2019 01:50:42

---

Diagnostics Check	OK	01:47	
Air Blank	0.000	01:48	
Control Test	0.079	01:48	
Air Blank	0.000	01:49	
Subject Sample	# 1	01:49	Subject Test Refused
Breath Volume	0.000	01:49	
Air Blank	0.000	01:49	
Control Test	0.078	01:50	
Air Blank	0.000	01:50	
Diagnostics Check	OK	01:50	

---

Data Download Date/Time: 09/03/2019 12:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARCIA,BRYAN,  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: THOMAS,M  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 07/26/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 640976  
DRY GAS EXPIRY DATE: 11/21/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: CASANOVA

DATE/TIME OF TEST: 08/04/2019 00:29:15

---

Diagnostics Check	OK	00:22
Air Blank	0.000	00:22
Control Test	0.078	00:22
Air Blank	0.000	00:23
Subject Sample # 1	0.142	00:24
Breath Volume	1.171	00:24
Air Blank	0.000	00:24
Air Blank	0.000	00:26
Subject Sample # 2	0.155	00:27
Breath Volume	1.746	00:27
Air Blank	0.000	00:28
Control Test	0.079	00:28
Air Blank	0.000	00:29
Diagnostics Check	OK	00:29

---

Data Download Date/Time: 09/03/2019 12:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WHITE,ROMI,T  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: HEREDIA RUBIO,ROBERTO  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 07/26/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 640976  
DRY GAS EXPIRY DATE: 11/21/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: BAEZ

DATE/TIME OF TEST: 08/09/2019 03:09:59

---

Diagnostics Check	OK	03:07	
Air Blank	0.000	03:07	
Control Test	0.078	03:07	
Air Blank	0.000	03:08	
Subject Sample	# 1	03:08	Subject Test Refused
Breath Volume	0.000	03:08	
Air Blank	0.000	03:09	
Control Test	0.078	03:09	
Air Blank	0.000	03:09	
Diagnostics Check	OK	03:09	

---

Data Download Date/Time: 09/03/2019 12:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAMIREZ,CRISTIAN,A  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: RAMIREZ,CRISTIAN  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 07/26/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 640976  
DRY GAS EXPIRY DATE: 11/21/2019  
VIOLATION CODE: DUI  
ARREST TIME: 11:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:40

SUBJECT'S LAST NAME: PENA-GONZALEZ

DATE/TIME OF TEST: 08/12/2019 12:04:25

---

Diagnostics Check	OK	12:01	
Air Blank	0.000	12:01	
Control Test	0.078	12:01	
Air Blank	0.000	12:02	
Subject Sample	# 1	12:03	Subject Test Refused
Breath Volume	0.000	12:03	
Air Blank	0.000	12:03	
Control Test	0.078	12:03	
Air Blank	0.000	12:04	
Diagnostics Check	OK	12:04	

---

Data Download Date/Time: 09/03/2019 12:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DEPRADO,DANIEL,L

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: DEPRADO,DANIEL

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 07/26/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 640976

DRY GAS EXPIRY DATE: 11/21/2019

VIOLATION CODE: DUI

ARREST TIME: 16:14

GENDER: Female

OBSERVATION PERIOD BEGAN: 17:33

SUBJECT'S LAST NAME: TUCKER GUMENICK

DATE/TIME OF TEST: 08/17/2019 18:08:08

---

Diagnostics Check	OK	17:58
Air Blank	0.000	17:58
Control Test	0.080	17:58
Air Blank	0.000	17:59
Subject Sample # 1	0.173	18:01
Breath Volume	1.902	18:01
Air Blank	0.000	18:02
Air Blank	0.000	18:03
Subject Sample # 2	0.173	18:06
Breath Volume	1.917	18:06
Air Blank	0.000	18:07
Control Test	0.076	18:07
Air Blank	0.000	18:08
Diagnostics Check	OK	18:08

---

Data Download Date/Time: 09/03/2019 12:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

INSTRUMENT REGISTERED TO: CALHOUN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUMPHREY, MATTHEW, D  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: BROWN, RAY  
ARRESTING OFFICER AGENCY: Calhoun County SO  
LAST AGENCY INSPECTION: 07/08/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29017080A3  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: SHULER

DATE/TIME OF TEST: 08/03/2019 22:55:39

---

Diagnostics Check	OK	22:49
Air Blank	0.000	22:49
Control Test	0.080	22:49
Air Blank	0.000	22:50
Subject Sample # 1	0.055	22:51
Breath Volume	2.902	22:51
Air Blank	0.000	22:51
Air Blank	0.000	22:53
Subject Sample # 2	0.062	22:54
Breath Volume	3.601	22:54
Air Blank	0.000	22:54
Control Test	0.080	22:55
Air Blank	0.000	22:55
Diagnostics Check	OK	22:55

---

Data Download Date/Time: 08/13/2019 07:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

INSTRUMENT REGISTERED TO: CALHOUN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMPSON,RONDRICKA,N  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: JACOBS,DAVID  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 08/12/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29017080A3  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 20:10  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: ROSIER

DATE/TIME OF TEST: 08/27/2019 21:31:55

---

Diagnostics Check		OK	21:20	
Air Blank		0.000	21:20	
Control Test		0.080	21:21	
Air Blank		0.000	21:21	
Subject Sample	# 1	0.079	21:24	Volume Not Met
Breath Volume		0.875	21:24	
Air Blank		0.000	21:25	
Air Blank		0.000	21:27	
Subject Sample	# 2	0.075	21:30	Volume Not Met
Breath Volume		0.960	21:30	
Air Blank		0.000	21:31	
Control Test		0.079	21:31	
Air Blank		0.000	21:31	
Diagnostics Check		OK	21:31	

---

Data Download Date/Time: 09/03/2019 09:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

INSTRUMENT REGISTERED TO: HOMESTEAD PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ,RAUL,  
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD  
ARRESTING OFFICER: BOHNE,STEVEN  
ARRESTING OFFICER AGENCY: HOMESTEAD PD  
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1114319  
DRY GAS EXPIRY DATE: 03/27/2022  
VIOLATION CODE: DUI  
ARREST TIME: 11:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:25

SUBJECT'S LAST NAME: RICHTER

DATE/TIME OF TEST: 08/10/2019 11:57:06

---

Diagnostics Check		OK	11:49
Air Blank		0.000	11:49
Control Test		0.082	11:50
Air Blank		0.000	11:50
Subject Sample	# 1	0.299	11:52
Breath Volume		1.363	11:52
Air Blank		0.000	11:53
Air Blank		0.000	11:54
Subject Sample	# 2	0.306	11:55
Breath Volume		1.304	11:55
Air Blank		0.000	11:56
Control Test		0.081	11:56
Air Blank		0.000	11:57
Diagnostics Check		OK	11:57

---

Data Download Date/Time: 08/30/2019 03:18



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CONFESSORE,STEVE,  
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD  
ARRESTING OFFICER: COSNER,STEVE  
ARRESTING OFFICER AGENCY: MIAMI BEACH PD  
LAST AGENCY INSPECTION: 07/31/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814181  
DRY GAS EXPIRY DATE: 07/24/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: HAREL

DATE/TIME OF TEST: 08/01/2019 00:43:11

---

Diagnostics Check	OK	00:40	
Air Blank	0.000	00:40	
Control Test	0.079	00:41	
Air Blank	0.000	00:41	
Subject Sample	# 1	00:41	Subject Test Refused
Breath Volume	0.000	00:41	
Air Blank	0.000	00:42	
Control Test	0.081	00:42	
Air Blank	0.000	00:43	
Diagnostics Check	OK	00:43	

---

Data Download Date/Time: 09/03/2019 15:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SOSA, JONATHAN,  
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD  
ARRESTING OFFICER: JIMENEZ, ANTHONY  
ARRESTING OFFICER AGENCY: NORTH BAY VILLAGE PD  
LAST AGENCY INSPECTION: 07/31/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814181  
DRY GAS EXPIRY DATE: 07/24/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: PAOLO

DATE/TIME OF TEST: 08/04/2019 02:19:54

---

Diagnostics Check	OK	02:11
Air Blank	0.000	02:11
Control Test	0.081	02:12
Air Blank	0.000	02:12
Subject Sample # 1	0.200	02:15
Breath Volume	1.562	02:15
Air Blank	0.000	02:16
Air Blank	0.000	02:17
Subject Sample # 2	0.216	02:18
Breath Volume	2.089	02:18
Air Blank	0.000	02:19
Control Test	0.081	02:19
Air Blank	0.000	02:19
Diagnostics Check	OK	02:19

---

Data Download Date/Time: 09/03/2019 15:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SOSA, JONATHAN,  
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD  
ARRESTING OFFICER: SOSA, JONATHAN  
ARRESTING OFFICER AGENCY: MIAMI BEACH PD  
LAST AGENCY INSPECTION: 07/31/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814181  
DRY GAS EXPIRY DATE: 07/24/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: MACHUCA

DATE/TIME OF TEST: 08/05/2019 04:59:05

---

Diagnostics Check	OK	04:54	
Air Blank	0.000	04:55	
Control Test	0.081	04:55	
Air Blank	0.000	04:55	
Subject Sample	# 1	04:57	Subject Test Refused
Breath Volume	0.000	04:57	
Air Blank	0.000	04:58	
Control Test	0.081	04:58	
Air Blank	0.000	04:59	
Diagnostics Check	OK	04:59	

---

Data Download Date/Time: 09/03/2019 15:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CASTILLO, RICARDO, A  
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD  
ARRESTING OFFICER: CASTILLO, RICARDO  
ARRESTING OFFICER AGENCY: MIAMI BEACH PD  
LAST AGENCY INSPECTION: 07/31/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814181  
DRY GAS EXPIRY DATE: 07/24/2020  
VIOLATION CODE: DUI  
ARREST TIME: 19:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: JOURDAIN

DATE/TIME OF TEST: 08/06/2019 19:54:30

---

Diagnostics Check	OK	19:44
Air Blank	0.000	19:44
Control Test	0.081	19:44
Air Blank	0.000	19:45
Subject Sample # 1	0.206	19:46
Breath Volume	2.976	19:46
Air Blank	0.000	19:46
Air Blank	0.000	19:48
Subject Sample # 2	0.175	19:49
Breath Volume	2.855	19:49
Air Blank	0.000	19:50
Air Blank	0.000	19:52
Subject Sample # 3	0.193	19:53
Breath Volume	2.250	19:53
Air Blank	0.000	19:53
Control Test	0.080	19:53
Air Blank	0.000	19:54
Diagnostics Check	OK	19:54

---

Data Download Date/Time: 09/03/2019 15:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PAIGE,LEON,I  
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD  
ARRESTING OFFICER: PAIGE,LEON  
ARRESTING OFFICER AGENCY: MIAMI BEACH PD  
LAST AGENCY INSPECTION: 07/31/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814181  
DRY GAS EXPIRY DATE: 07/24/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 08/10/2019 00:43:31

---

Diagnostics Check	OK	00:40	
Air Blank	0.000	00:41	
Control Test	0.081	00:41	
Air Blank	0.000	00:41	
Subject Sample	# 1	00:42	Subject Test Refused
Breath Volume	0.000	00:42	
Air Blank	0.000	00:42	
Control Test	0.081	00:42	
Air Blank	0.000	00:43	
Diagnostics Check	OK	00:43	

---

Data Download Date/Time: 09/03/2019 15:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CONFESSORE, STEVE,  
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD  
ARRESTING OFFICER: MELENDEZ, DANIEL  
ARRESTING OFFICER AGENCY: MIAMI BEACH PD  
LAST AGENCY INSPECTION: 07/31/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814181  
DRY GAS EXPIRY DATE: 07/24/2020  
VIOLATION CODE: DUI  
ARREST TIME: 04:36  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: GOODWIN

DATE/TIME OF TEST: 08/16/2019 05:24:19

---

Diagnostics Check	OK	05:12	
Air Blank	0.000	05:13	
Control Test	0.081	05:13	
Air Blank	0.000	05:14	
Subject Sample	# 1	05:17	No Sample Provided
Breath Volume	0.000	05:17	
Air Blank	0.000	05:17	
Air Blank	0.000	05:19	
Subject Sample	# 2	05:22	No Sample Provided
Breath Volume	0.000	05:22	
Air Blank	0.000	05:23	
Control Test	0.081	05:23	
Air Blank	0.000	05:24	
Diagnostics Check	OK	05:24	

---

Data Download Date/Time: 09/03/2019 15:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CONFESSORE,STEVE,  
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD  
ARRESTING OFFICER: MELENDEZ,DANIEL  
ARRESTING OFFICER AGENCY: MIAMI BEACH PD  
LAST AGENCY INSPECTION: 07/31/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814181  
DRY GAS EXPIRY DATE: 07/24/2020  
VIOLATION CODE: DUI  
ARREST TIME: 04:36  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: GOODWIN

DATE/TIME OF TEST: 08/16/2019 05:33:00

---

Diagnostics Check	OK	05:26
Air Blank	0.000	05:27
Control Test	0.081	05:27
Air Blank	0.000	05:28
Subject Sample # 1	0.185	05:28
Breath Volume	1.746	05:28
Air Blank	0.000	05:29
Air Blank	0.000	05:31
Subject Sample # 2	0.171	05:31
Breath Volume	2.339	05:31
Air Blank	0.000	05:32
Control Test	0.080	05:32
Air Blank	0.000	05:32
Diagnostics Check	OK	05:33

---

Data Download Date/Time: 09/03/2019 15:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SOSA, JONATHAN,  
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD  
ARRESTING OFFICER: SOSA, JONATHAN  
ARRESTING OFFICER AGENCY: MIAMI BEACH PD  
LAST AGENCY INSPECTION: 07/31/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814181  
DRY GAS EXPIRY DATE: 07/24/2020  
VIOLATION CODE: DUI  
ARREST TIME: 05:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:49

SUBJECT'S LAST NAME: LLUESMA VIDAL

DATE/TIME OF TEST: 08/17/2019 06:23:36

---

Diagnostics Check	OK	06:18	
Air Blank	0.000	06:19	
Control Test	0.081	06:19	
Air Blank	0.000	06:20	
Subject Sample	# 1	06:22	Subject Test Refused
Breath Volume	0.000	06:22	
Air Blank	0.000	06:22	
Control Test	0.080	06:23	
Air Blank	0.000	06:23	
Diagnostics Check	OK	06:23	

---

Data Download Date/Time: 09/03/2019 15:50



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CABANAS,ROBERT,  
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD  
ARRESTING OFFICER: JIMENEZ,ANTHONY  
ARRESTING OFFICER AGENCY: NORTH BAY VILLAGE PD  
LAST AGENCY INSPECTION: 07/31/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814181  
DRY GAS EXPIRY DATE: 07/24/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: CAPOTE

DATE/TIME OF TEST: 08/18/2019 03:27:53

---

Diagnostics Check		OK	03:21
Air Blank		0.000	03:22
Control Test		0.082	03:22
Air Blank		0.000	03:22
Subject Sample	# 1	0.000	03:23
Breath Volume		2.710	03:23
Air Blank		0.000	03:24
Air Blank		0.000	03:26
Subject Sample	# 2	0.000	03:26
Breath Volume		4.000	03:26
Air Blank		0.000	03:26
Control Test		0.081	03:27
Air Blank		0.000	03:27
Diagnostics Check		OK	03:27

---

Data Download Date/Time: 09/03/2019 15:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SOSA, JONATHAN,  
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD  
ARRESTING OFFICER: SOSA, SOSA  
ARRESTING OFFICER AGENCY: MIAMI BEACH PD  
LAST AGENCY INSPECTION: 07/31/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814181  
DRY GAS EXPIRY DATE: 07/24/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: GUZMAN

DATE/TIME OF TEST: 08/24/2019 03:07:58

---

Diagnostics Check	OK	02:58	
Air Blank	0.000	02:58	
Control Test	0.081	02:58	
Air Blank	0.000	02:59	
Subject Sample	# 1	03:02	No Sample Provided
Breath Volume	0.000	03:02	
Air Blank	0.000	03:03	
Air Blank	0.000	03:05	
Subject Sample	# 2	03:06	Subject Test Refused
Breath Volume	0.000	03:06	
Air Blank	0.000	03:07	
Control Test	0.082	03:07	
Air Blank	0.000	03:07	
Diagnostics Check	OK	03:07	

---

Data Download Date/Time: 09/03/2019 15:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PAIGE,LEON,I  
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD  
ARRESTING OFFICER: PAIGE,LEON  
ARRESTING OFFICER AGENCY: MIAMI BEACH PD  
LAST AGENCY INSPECTION: 07/31/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814181  
DRY GAS EXPIRY DATE: 07/24/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: FLORES-ZAVALA

DATE/TIME OF TEST: 08/25/2019 04:57:41

---

Diagnostics Check	OK	04:50
Air Blank	0.000	04:50
Control Test	0.081	04:50
Air Blank	0.000	04:51
Subject Sample # 1	0.198	04:52
Breath Volume	2.109	04:52
Air Blank	0.000	04:53
Air Blank	0.000	04:54
Subject Sample # 2	0.196	04:56
Breath Volume	3.109	04:56
Air Blank	0.000	04:56
Control Test	0.081	04:57
Air Blank	0.000	04:57
Diagnostics Check	OK	04:57

---

Data Download Date/Time: 09/03/2019 15:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SOSA, JONATHAN,  
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD  
ARRESTING OFFICER: VEGODA, JOSHUA  
ARRESTING OFFICER AGENCY: MIAMI BEACH PD  
LAST AGENCY INSPECTION: 07/31/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814181  
DRY GAS EXPIRY DATE: 07/24/2020  
VIOLATION CODE: DUI  
ARREST TIME: 04:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:54

SUBJECT'S LAST NAME: VILLAR

DATE/TIME OF TEST: 08/25/2019 05:23:08

---

Diagnostics Check	OK	05:15
Air Blank	0.000	05:15
Control Test	0.080	05:15
Air Blank	0.000	05:16
Subject Sample # 1	0.123	05:18
Breath Volume	3.460	05:18
Air Blank	0.000	05:19
Air Blank	0.000	05:21
Subject Sample # 2	0.134	05:21
Breath Volume	2.000	05:21
Air Blank	0.000	05:22
Control Test	0.081	05:22
Air Blank	0.000	05:23
Diagnostics Check	OK	05:23

---

Data Download Date/Time: 09/03/2019 15:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRIND,ABEL,D  
BREATH TEST OPERATOR AGENCY: NORTH BAY VILLAGE PD  
ARRESTING OFFICER: BRIND,ABEL  
ARRESTING OFFICER AGENCY: NORTH BAY VILLAGE PD  
LAST AGENCY INSPECTION: 07/31/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814181  
DRY GAS EXPIRY DATE: 07/24/2020  
VIOLATION CODE: DUI  
ARREST TIME: 06:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:48

SUBJECT'S LAST NAME: TIRADO GARCIA  
DATE/TIME OF TEST: 08/25/2019 07:16:44

---

Diagnostics Check	OK	07:09
Air Blank	0.000	07:10
Control Test	0.081	07:10
Air Blank	0.000	07:10
Subject Sample # 1	0.230	07:11
Breath Volume	1.316	07:11
Air Blank	0.000	07:12
Air Blank	0.000	07:14
Subject Sample # 2	0.227	07:15
Breath Volume	2.421	07:15
Air Blank	0.000	07:15
Control Test	0.081	07:16
Air Blank	0.000	07:16
Diagnostics Check	OK	07:16

---

Data Download Date/Time: 09/03/2019 15:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DOMINGUEZ,FREDERICK,A  
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD  
ARRESTING OFFICER: DOMINGUEZ,FREDERICK  
ARRESTING OFFICER AGENCY: MIAMI BEACH PD  
LAST AGENCY INSPECTION: 07/31/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814181  
DRY GAS EXPIRY DATE: 07/24/2020  
VIOLATION CODE: DUI  
ARREST TIME: 09:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:56

SUBJECT'S LAST NAME: ZORGANI

DATE/TIME OF TEST: 08/25/2019 10:22:54

---

Diagnostics Check	OK	10:16
Air Blank	0.000	10:16
Control Test	0.082	10:17
Air Blank	0.000	10:17
Subject Sample # 1	0.139	10:18
Breath Volume	2.246	10:18
Air Blank	0.000	10:18
Air Blank	0.000	10:20
Subject Sample # 2	0.141	10:21
Breath Volume	2.230	10:21
Air Blank	0.000	10:22
Control Test	0.081	10:22
Air Blank	0.000	10:22
Diagnostics Check	OK	10:22

---

Data Download Date/Time: 09/03/2019 15:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SOCARRAS,PEDRO,A  
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD  
ARRESTING OFFICER: PAIGE,LEON  
ARRESTING OFFICER AGENCY: MIAMI BEACH PD  
LAST AGENCY INSPECTION: 07/31/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814181  
DRY GAS EXPIRY DATE: 07/24/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 08/29/2019 02:02:12

---

Diagnostics Check	OK	01:55
Air Blank	0.000	01:56
Control Test	0.082	01:56
Air Blank	0.000	01:56
Subject Sample # 1	0.192	01:57
Breath Volume	1.539	01:57
Air Blank	0.000	01:58
Air Blank	0.000	02:00
Subject Sample # 2	0.189	02:00
Breath Volume	1.324	02:00
Air Blank	0.000	02:01
Control Test	0.081	02:01
Air Blank	0.000	02:02
Diagnostics Check	OK	02:02

---

Data Download Date/Time: 09/03/2019 15:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRIND,ABEL,D  
BREATH TEST OPERATOR AGENCY: NORTH BAY VILLAGE PD  
ARRESTING OFFICER: CHEVALIER,ISMAEL  
ARRESTING OFFICER AGENCY: NORTH BAY VILLAGE PD  
LAST AGENCY INSPECTION: 07/31/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814181  
DRY GAS EXPIRY DATE: 07/24/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: GREY  
DATE/TIME OF TEST: 08/29/2019 04:04:50

---

Diagnostics Check	OK	03:58
Air Blank	0.000	03:58
Control Test	0.081	03:58
Air Blank	0.000	03:59
Subject Sample # 1	0.240	04:00
Breath Volume	1.718	04:00
Air Blank	0.000	04:00
Air Blank	0.000	04:02
Subject Sample # 2	0.250	04:03
Breath Volume	2.257	04:03
Air Blank	0.000	04:03
Control Test	0.081	04:04
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

---

Data Download Date/Time: 09/03/2019 15:51



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PACHECO,GEORGE,  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: DELGADO,M  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 08/08/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 06:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:30

SUBJECT'S LAST NAME: DIAZ DIAZ

DATE/TIME OF TEST: 08/09/2019 07:03:14

---

Diagnostics Check	OK	06:56
Air Blank	0.000	06:56
Control Test	0.080	06:56
Air Blank	0.000	06:57
Subject Sample # 1	0.176	06:58
Breath Volume	2.281	06:58
Air Blank	0.000	06:58
Air Blank	0.000	07:00
Subject Sample # 2	0.169	07:01
Breath Volume	1.640	07:01
Air Blank	0.000	07:02
Control Test	0.079	07:02
Air Blank	0.000	07:03
Diagnostics Check	OK	07:03

---

Data Download Date/Time: 09/04/2019 09:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROIG,DAWN,M  
BREATH TEST OPERATOR AGENCY: MIAMI SHORES PD  
ARRESTING OFFICER: ROIG,DAWN  
ARRESTING OFFICER AGENCY: MIAMI SHORES PD  
LAST AGENCY INSPECTION: 08/08/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:49

SUBJECT'S LAST NAME: OSIAS

DATE/TIME OF TEST: 08/11/2019 21:15:47

---

Diagnostics Check		OK	21:09
Air Blank		0.000	21:10
Control Test		0.079	21:10
Air Blank		0.000	21:10
Subject Sample	# 1	0.236	21:11
Breath Volume		1.324	21:11
Air Blank		0.000	21:11
Air Blank		0.000	21:13
Subject Sample	# 2	0.244	21:14
Breath Volume		1.925	21:14
Air Blank		0.000	21:14
Control Test		0.080	21:15
Air Blank		0.000	21:15
Diagnostics Check		OK	21:15

---

Data Download Date/Time: 09/04/2019 09:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LOPEZ,KEVIN,M  
BREATH TEST OPERATOR AGENCY: OTHER LOCAL  
ARRESTING OFFICER: LOPEZ,KEVIN  
ARRESTING OFFICER AGENCY: OTHER COUNTY  
LAST AGENCY INSPECTION: 08/08/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 15:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:15

SUBJECT'S LAST NAME: MCDONALD

DATE/TIME OF TEST: 08/12/2019 16:55:06

---

Diagnostics Check	OK	16:52	
Air Blank	0.000	16:52	
Control Test	0.079	16:52	
Air Blank	0.000	16:53	
Subject Sample	# 1	16:53	Subject Test Refused
Breath Volume	0.000	16:53	
Air Blank	0.000	16:54	
Control Test	0.078	16:54	
Air Blank	0.000	16:55	
Diagnostics Check	OK	16:55	

---

Data Download Date/Time: 09/04/2019 09:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HELFNER,ISAAC,S  
BREATH TEST OPERATOR AGENCY: AVENTURA PD  
ARRESTING OFFICER: PEREZ,EMILIO  
ARRESTING OFFICER AGENCY: AVENTURA PD  
LAST AGENCY INSPECTION: 08/08/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: MORENO

DATE/TIME OF TEST: 08/27/2019 23:25:48

---

Diagnostics Check	OK	23:19
Air Blank	0.000	23:20
Control Test	0.077	23:20
Air Blank	0.000	23:20
Subject Sample # 1	0.329	23:21
Breath Volume	1.398	23:21
Air Blank	0.000	23:22
Air Blank	0.000	23:23
Subject Sample # 2	0.312	23:24
Breath Volume	1.906	23:24
Air Blank	0.000	23:24
Control Test	0.078	23:25
Air Blank	0.000	23:25
Diagnostics Check	OK	23:25

---

Data Download Date/Time: 09/04/2019 09:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GABRIEL, EVENS,  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: DELGADO, MIGUEL  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 08/08/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: PAUL

DATE/TIME OF TEST: 08/29/2019 03:44:42

---

Diagnostics Check	OK	03:41	
Air Blank	0.000	03:41	
Control Test	0.081	03:41	
Air Blank	0.000	03:42	
Subject Sample	# 1	03:43	Slope Not Met
Breath Volume	0.628	03:43	
Air Blank	0.000	03:43	
Control Test	0.081	03:44	
Air Blank	0.000	03:44	
Diagnostics Check	OK	03:44	

---

Data Download Date/Time: 09/04/2019 09:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GABRIEL, EVENS,  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: DELGADO, MIGUEL  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 08/08/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: PAUL

DATE/TIME OF TEST: 08/29/2019 03:59:01

---

Diagnostics Check		OK	03:46	
Air Blank		0.000	03:46	
Control Test		0.079	03:46	
Air Blank		0.000	03:47	
Subject Sample	# 1	0.188	03:50	Volume Not Met
Breath Volume		0.730	03:50	
Air Blank		0.000	03:51	
Air Blank		0.000	03:53	
Subject Sample	# 2	0.196	03:54	
Breath Volume		1.167	03:54	
Air Blank		0.000	03:55	
Air Blank		0.000	03:56	
Subject Sample	# 3	0.194	03:57	
Breath Volume		1.570	03:57	
Air Blank		0.000	03:58	
Control Test		0.076	03:58	
Air Blank		0.000	03:58	
Diagnostics Check		OK	03:59	

---

Data Download Date/Time: 09/04/2019 09:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROMERO,JANET,  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: ROMERO,JANET  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 08/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34917080A4  
DRY GAS EXPIRY DATE: 02/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: DEJESUS

DATE/TIME OF TEST: 08/09/2019 02:40:53

---

Diagnostics Check	OK	02:37	
Air Blank	0.000	02:38	
Control Test	0.079	02:38	
Air Blank	0.000	02:39	
Subject Sample	# 1	02:39	Subject Test Refused
Breath Volume	0.000	02:39	
Air Blank	0.000	02:39	
Control Test	0.079	02:40	
Air Blank	0.000	02:40	
Diagnostics Check	OK	02:40	

---

Data Download Date/Time: 09/17/2019 08:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BONILLA,JOSE,I  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: BONILLA,JOSE  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 08/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34917080A4  
DRY GAS EXPIRY DATE: 02/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: MACHADO

DATE/TIME OF TEST: 08/15/2019 02:41:19

---

Diagnostics Check	OK	02:38	
Air Blank	0.000	02:38	
Control Test	0.079	02:39	
Air Blank	0.000	02:39	
Subject Sample	# 1	02:39	Subject Test Refused
Breath Volume	0.000	02:39	
Air Blank	0.000	02:40	
Control Test	0.079	02:40	
Air Blank	0.000	02:41	
Diagnostics Check	OK	02:41	

---

Data Download Date/Time: 09/17/2019 08:10



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PHILPOT,C,  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: PHILPOT,C  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 08/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34917080A4  
DRY GAS EXPIRY DATE: 02/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: ZAYAS

DATE/TIME OF TEST: 08/17/2019 23:26:58

---

Diagnostics Check	OK	23:20
Air Blank	0.000	23:21
Control Test	0.079	23:21
Air Blank	0.000	23:22
Subject Sample # 1	0.073	23:22
Breath Volume	1.613	23:22
Air Blank	0.000	23:23
Air Blank	0.000	23:25
Subject Sample # 2	0.068	23:25
Breath Volume	2.472	23:25
Air Blank	0.000	23:26
Control Test	0.079	23:26
Air Blank	0.000	23:26
Diagnostics Check	OK	23:26

---

Data Download Date/Time: 09/17/2019 08:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HERNANDEZ,KEITH,C  
BREATH TEST OPERATOR AGENCY: FFWCC (SOUTH A)  
ARRESTING OFFICER: HERNANDEZ,KEITH  
ARRESTING OFFICER AGENCY: FFWCC (SOUTH A)  
LAST AGENCY INSPECTION: 08/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34917080A4  
DRY GAS EXPIRY DATE: 02/05/2020  
VIOLATION CODE: BUI  
ARREST TIME: 23:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: PENDER

DATE/TIME OF TEST: 08/18/2019 01:23:29

---

Diagnostics Check	OK	01:19	
Air Blank	0.000	01:20	
Control Test	0.079	01:20	
Air Blank	0.000	01:20	
Subject Sample	# 1	01:21	RFI Detect
Breath Volume	0.199	01:21	
Air Blank	0.000	01:22	
Control Test	0.079	01:22	
Air Blank	0.000	01:23	
Diagnostics Check	OK	01:23	

---

Data Download Date/Time: 09/17/2019 08:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HERNANDEZ,KEITH,C  
BREATH TEST OPERATOR AGENCY: FFWCC (SOUTH A)  
ARRESTING OFFICER: HERNANDEZ,KEITH  
ARRESTING OFFICER AGENCY: FFWCC (SOUTH A)  
LAST AGENCY INSPECTION: 08/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34917080A4  
DRY GAS EXPIRY DATE: 02/05/2020  
VIOLATION CODE: BUI  
ARREST TIME: 23:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: PENDER

DATE/TIME OF TEST: 08/18/2019 01:39:10

---

Diagnostics Check	OK	01:30
Air Blank	0.000	01:30
Control Test	0.078	01:30
Air Blank	0.000	01:31
Subject Sample # 1	0.163	01:32
Breath Volume	1.183	01:32
Air Blank	0.000	01:33
Air Blank	0.000	01:34
Subject Sample # 2	0.180	01:37
Breath Volume	1.226	01:37
Air Blank	0.000	01:38
Control Test	0.078	01:38
Air Blank	0.000	01:39
Diagnostics Check	OK	01:39

---

Data Download Date/Time: 09/17/2019 08:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SLIMAK,M,  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: SLIMAK,M  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 08/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34917080A4  
DRY GAS EXPIRY DATE: 02/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: ROQUE VASQUEZ

DATE/TIME OF TEST: 08/18/2019 04:31:35

---

Diagnostics Check	OK	04:25
Air Blank	0.000	04:25
Control Test	0.077	04:26
Air Blank	0.000	04:26
Subject Sample # 1	0.240	04:27
Breath Volume	2.328	04:27
Air Blank	0.000	04:27
Air Blank	0.000	04:29
Subject Sample # 2	0.234	04:29
Breath Volume	1.843	04:29
Air Blank	0.000	04:30
Control Test	0.077	04:30
Air Blank	0.000	04:31
Diagnostics Check	OK	04:31

---

Data Download Date/Time: 09/17/2019 08:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GILPIN,SCHELISA,  
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD  
ARRESTING OFFICER: GILPIN,SCHELISA  
ARRESTING OFFICER AGENCY: HOMESTEAD PD  
LAST AGENCY INSPECTION: 08/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34917080A4  
DRY GAS EXPIRY DATE: 02/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 08/30/2019 03:15:35

---

Diagnostics Check	OK	03:08
Air Blank	0.000	03:08
Control Test	0.079	03:09
Air Blank	0.000	03:09
Subject Sample # 1	0.107	03:10
Breath Volume	1.359	03:10
Air Blank	0.000	03:11
Air Blank	0.000	03:13
Subject Sample # 2	0.115	03:13
Breath Volume	1.953	03:13
Air Blank	0.000	03:14
Control Test	0.077	03:14
Air Blank	0.000	03:15
Diagnostics Check	OK	03:15

---

Data Download Date/Time: 09/17/2019 08:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEANDRE,GERALD,L  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: LEANDRE,GERALD  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 08/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34917080A4  
DRY GAS EXPIRY DATE: 02/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: HENDERSON

DATE/TIME OF TEST: 08/31/2019 01:48:28

---

Diagnostics Check	OK	01:41	
Air Blank	0.000	01:41	
Control Test	0.079	01:42	
Air Blank	0.000	01:42	
Subject Sample # 1	0.224	01:43	
Breath Volume	2.191	01:43	
Air Blank	0.000	01:44	
Air Blank	0.000	01:46	
Subject Sample # 2	0.198	01:46	
Breath Volume	3.597	01:46	
Air Blank		01:46	Improper Sample
Air Blank	0.000	01:47	
Control Test	0.076	01:47	
Air Blank	0.000	01:48	
Diagnostics Check	OK	01:48	

---

Data Download Date/Time: 09/17/2019 08:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEANDRE,GERALD,L  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: LEANDRE,GERALD  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 08/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34917080A4  
DRY GAS EXPIRY DATE: 02/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: HENDERSON

DATE/TIME OF TEST: 08/31/2019 01:56:40

---

Diagnostics Check	OK	01:52	
Air Blank	0.000	01:52	
Control Test	0.076	01:52	
Air Blank	0.000	01:53	
Subject Sample # 1	0.223	01:54	
Breath Volume	1.820	01:54	
Air Blank		01:55	Purge Fail
Air Blank	0.000	01:56	

---

Data Download Date/Time: 09/17/2019 08:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEANDRE,GERALD,L  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: LEANDRE,GERALD  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 08/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34917080A4  
DRY GAS EXPIRY DATE: 02/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: HENDERSON

DATE/TIME OF TEST: 08/31/2019 02:01:11

---

Diagnostics Check	OK	01:59	
Air Blank	0.000	02:00	
Control Test	0.073	02:00	Control Outside Tolerance
Air Blank	0.000	02:01	

---

Data Download Date/Time: 09/17/2019 08:11



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VILLANUEVA, MAURICIO,  
BREATH TEST OPERATOR AGENCY: North Miami Beach PD  
ARRESTING OFFICER: ELIAS, NATALIE  
ARRESTING OFFICER AGENCY: North Miami Beach PD  
LAST AGENCY INSPECTION: 07/31/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34917080A4  
DRY GAS EXPIRY DATE: 02/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: LAINE

DATE/TIME OF TEST: 08/18/2019 02:05:09

---

Diagnostics Check	OK	01:58
Air Blank	0.000	01:58
Control Test	0.079	01:59
Air Blank	0.000	01:59
Subject Sample # 1	0.173	02:00
Breath Volume	1.476	02:00
Air Blank	0.000	02:00
Air Blank	0.000	02:02
Subject Sample # 2	0.170	02:03
Breath Volume	1.535	02:03
Air Blank	0.000	02:04
Control Test	0.078	02:04
Air Blank	0.000	02:05
Diagnostics Check	OK	02:05

---

Data Download Date/Time: 08/30/2019 13:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VILLANUEVA, MAURICIO,  
BREATH TEST OPERATOR AGENCY: North Miami Beach PD  
ARRESTING OFFICER: VILLANUEVA, MAURICIO  
ARRESTING OFFICER AGENCY: North Miami Beach PD  
LAST AGENCY INSPECTION: 07/31/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34917080A4  
DRY GAS EXPIRY DATE: 02/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 04:09  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: BIRMAN

DATE/TIME OF TEST: 08/18/2019 04:49:29

---

Diagnostics Check	OK	04:40	
Air Blank	0.000	04:41	
Control Test	0.078	04:41	
Air Blank	0.000	04:42	
Subject Sample	# 1	04:45	No Sample Provided
Breath Volume	0.000	04:45	
Air Blank	0.000	04:45	
Air Blank	0.000	04:47	
Subject Sample	# 2	04:48	Subject Test Refused
Breath Volume	0.000	04:48	
Air Blank	0.000	04:48	
Control Test	0.079	04:48	
Air Blank	0.000	04:49	
Diagnostics Check	OK	04:49	

---

Data Download Date/Time: 08/30/2019 13:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MAYA,JOSE,  
BREATH TEST OPERATOR AGENCY: North Miami Beach PD  
ARRESTING OFFICER: ALADIN,KENDRICK  
ARRESTING OFFICER AGENCY: North Miami Beach PD  
LAST AGENCY INSPECTION: 07/31/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34917080A4  
DRY GAS EXPIRY DATE: 02/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: MOCK

DATE/TIME OF TEST: 08/30/2019 00:37:10

---

Diagnostics Check	OK	00:30
Air Blank	0.000	00:30
Control Test	0.077	00:31
Air Blank	0.000	00:31
Subject Sample # 1	0.189	00:32
Breath Volume	2.109	00:32
Air Blank	0.000	00:33
Air Blank	0.000	00:34
Subject Sample # 2	0.202	00:35
Breath Volume	1.257	00:35
Air Blank	0.000	00:36
Control Test	0.077	00:36
Air Blank	0.000	00:37
Diagnostics Check	OK	00:37

---

Data Download Date/Time: 08/30/2019 13:42

**Florida Department of Law Enforcement**

**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILKINSON,WILLIAM,T  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: MCVANEY,JEFFERY  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 07/17/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 962587  
DRY GAS EXPIRY DATE: 06/04/2021  
VIOLATION CODE: BUI  
ARREST TIME: 13:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:00

SUBJECT'S LAST NAME: HIGLEY

DATE/TIME OF TEST: 08/03/2019 13:38:36

---

Diagnostics Check	OK	13:30	
Air Blank	0.000	13:30	
Control Test	0.080	13:30	
Air Blank	0.000	13:31	
Subject Sample	# 1	13:34	No Sample Provided
Breath Volume	0.000	13:34	
Air Blank	0.000	13:35	
Air Blank	0.000	13:37	
Subject Sample	# 2	13:37	Subject Test Refused
Breath Volume	0.000	13:37	
Air Blank	0.000	13:37	
Control Test	0.080	13:38	
Air Blank	0.000	13:38	
Diagnostics Check	OK	13:38	

---

Data Download Date/Time: 08/22/2019 13:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAMOS,ROBERT,  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: CORBIN,BRITTON  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 07/17/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 962587  
DRY GAS EXPIRY DATE: 06/04/2021  
VIOLATION CODE: BUI  
ARREST TIME: 16:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:59

SUBJECT'S LAST NAME: BATES

DATE/TIME OF TEST: 08/03/2019 17:48:05

---

Diagnostics Check	OK	17:36	
Air Blank	0.000	17:37	
Control Test	0.080	17:37	
Air Blank	0.000	17:37	
Subject Sample	# 1	17:41	No Sample Provided
Breath Volume	0.000	17:41	
Air Blank	0.000	17:41	
Air Blank	0.000	17:43	
Subject Sample	# 2	17:46	No Sample Provided
Breath Volume	0.000	17:46	
Air Blank	0.000	17:47	
Control Test	0.080	17:47	
Air Blank	0.000	17:48	
Diagnostics Check	OK	17:48	

---

Data Download Date/Time: 08/22/2019 13:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARTLETT,SHELTON,T  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: MOLNAR,JARROD  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 07/17/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 962587  
DRY GAS EXPIRY DATE: 06/04/2021  
VIOLATION CODE: BUI  
ARREST TIME: 18:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: ARCENEUX

DATE/TIME OF TEST: 08/03/2019 18:33:26

---

Diagnostics Check	OK	18:27
Air Blank	0.000	18:27
Control Test	0.080	18:27
Air Blank	0.000	18:28
Subject Sample # 1	0.192	18:29
Breath Volume	4.394	18:29
Air Blank	0.000	18:29
Air Blank	0.000	18:31
Subject Sample # 2	0.190	18:31
Breath Volume	4.304	18:31
Air Blank	0.000	18:32
Control Test	0.077	18:32
Air Blank	0.000	18:33
Diagnostics Check	OK	18:33

---

Data Download Date/Time: 08/22/2019 13:54



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CORBIN,BRITTON,L  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: CORBIN,BRITTON  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 07/17/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 962587  
DRY GAS EXPIRY DATE: 06/04/2021  
VIOLATION CODE: BUI  
ARREST TIME: 17:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: HUISKEN

DATE/TIME OF TEST: 08/11/2019 18:07:34

---

Diagnostics Check	OK	17:58	
Air Blank	0.000	17:59	
Control Test	0.080	17:59	
Air Blank	0.000	17:59	
Subject Sample	# 1	18:02	No Sample Provided
Breath Volume	0.000	18:02	
Air Blank	0.000	18:03	
Air Blank	0.000	18:05	
Subject Sample	# 2	18:06	Subject Test Refused
Breath Volume	0.000	18:06	
Air Blank	0.000	18:06	
Control Test	0.081	18:06	
Air Blank	0.000	18:07	
Diagnostics Check	OK	18:07	

---

Data Download Date/Time: 08/22/2019 13:54



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CORBITT,KYLE,L  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: WAGNER,ROBERT  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 07/17/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 962587  
DRY GAS EXPIRY DATE: 06/04/2021  
VIOLATION CODE: BUI  
ARREST TIME: 17:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:37

SUBJECT'S LAST NAME: SPEAR

DATE/TIME OF TEST: 08/11/2019 18:34:33

---

Diagnostics Check	OK	18:21	
Air Blank	0.000	18:21	
Control Test	0.080	18:21	
Air Blank	0.000	18:22	
Subject Sample	# 1	18:25	No Sample Provided
Breath Volume	0.000	18:25	
Air Blank	0.000	18:26	
Air Blank	0.000	18:28	
Subject Sample	# 2	0.141	18:29
Breath Volume	2.582	18:29	
Air Blank	0.000	18:30	
Air Blank	0.000	18:31	
Subject Sample	# 3	0.126	18:32
Breath Volume	2.246	18:32	
Air Blank	0.000	18:33	
Control Test	0.081	18:33	
Air Blank	0.000	18:34	
Diagnostics Check	OK	18:34	

---

Data Download Date/Time: 08/22/2019 13:54

**Florida Department of Law Enforcement**

**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NELSON,THOMAS,E  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: CORBIN,BRITTON  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 07/17/2019

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 962587  
DRY GAS EXPIRY DATE: 06/04/2021  
VIOLATION CODE: BUI  
ARREST TIME: 18:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: LEGARE

DATE/TIME OF TEST: 08/16/2019 19:35:56

---

Diagnostics Check	OK	19:27	
Air Blank	0.000	19:27	
Control Test	0.079	19:28	
Air Blank	0.000	19:28	
Subject Sample	# 1	19:31	No Sample Provided
Breath Volume	0.000	19:31	
Air Blank	0.000	19:32	
Air Blank	0.000	19:34	
Subject Sample	# 2	19:34	Subject Test Refused
Breath Volume	0.000	19:34	
Air Blank	0.000	19:35	
Control Test	0.079	19:35	
Air Blank	0.000	19:35	
Diagnostics Check	OK	19:35	

---

Data Download Date/Time: 08/22/2019 13:54

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 962587
BREATH TEST OPERATOR: CORBIN,BRITTON,L	DRY GAS EXPIRY DATE: 06/04/2021
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)	VIOLATION CODE: BUI
ARRESTING OFFICER: CORBIN,BRITTON	ARREST TIME: 16:48
ARRESTING OFFICER AGENCY: FFWCC (Panama City)	GENDER: Male
LAST AGENCY INSPECTION: 07/17/2019	OBSERVATION PERIOD BEGAN: 16:48

SUBJECT'S LAST NAME: WORLEY	DATE/TIME OF TEST: 08/17/2019	17:17:37
-----------------------------	-------------------------------	----------

---

Diagnostics Check	OK	17:08	
Air Blank	0.000	17:09	
Control Test	0.079	17:09	
Air Blank	0.000	17:10	
Subject Sample # 1		17:13	No Sample Provided
Breath Volume	0.000	17:13	
Air Blank	0.000	17:13	
Air Blank	0.000	17:15	
Subject Sample # 2		17:16	Subject Test Refused
Breath Volume	0.000	17:16	
Air Blank	0.000	17:16	
Control Test	0.080	17:17	
Air Blank	0.000	17:17	
Diagnostics Check	OK	17:17	

---

Data Download Date/Time: 08/22/2019 13:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WAGNER,ROBERT,A  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: CULBERSON,DARYL  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 07/17/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 962587  
DRY GAS EXPIRY DATE: 06/04/2021  
VIOLATION CODE: BUI  
ARREST TIME: 17:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:07

SUBJECT'S LAST NAME: BREAZEALE

DATE/TIME OF TEST: 08/17/2019 17:57:03

---

Diagnostics Check	OK	17:48	
Air Blank	0.000	17:48	
Control Test	0.080	17:49	
Air Blank	0.000	17:49	
Subject Sample	# 1	17:52	No Sample Provided
Breath Volume	0.000	17:52	
Air Blank	0.000	17:53	
Air Blank	0.000	17:55	
Subject Sample	# 2	17:55	Subject Test Refused
Breath Volume	0.000	17:55	
Air Blank	0.000	17:56	
Control Test	0.080	17:56	
Air Blank	0.000	17:56	
Diagnostics Check	OK	17:57	

---

Data Download Date/Time: 08/22/2019 13:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC  
 SOFTWARE VERSION: 8100.27  
 BREATH TEST OPERATOR: MALTAIS,ANDREW,W  
 BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
 ARRESTING OFFICER: MALTAIS,ANDREW  
 ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
 LAST AGENCY INSPECTION: 08/22/2019

Date or Time Changed: No  
 DRY GAS STANDARD LOT #: 962587  
 DRY GAS EXPIRY DATE: 06/04/2021  
 VIOLATION CODE: DUI  
 ARREST TIME: 19:29  
 GENDER: Male  
 OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: FAILS

DATE/TIME OF TEST: 08/25/2019 19:15:28

Diagnostics Check	OK	19:09
Air Blank	0.000	19:09
Control Test	0.080	19:10
Air Blank	0.000	19:10
Subject Sample # 1	0.165	19:11
Breath Volume	3.093	19:11
Air Blank	0.000	19:11
Air Blank	0.000	19:13
Subject Sample # 2	0.158	19:13
Breath Volume	4.152	19:13
Air Blank	0.000	19:14
Control Test	0.079	19:14
Air Blank	0.000	19:15
Diagnostics Check	OK	19:15

Data Download Date/Time: 09/11/2019 16:29



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CORBITT,KYLE,L  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: WAGNER,ROBERT  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 08/22/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 962587  
DRY GAS EXPIRY DATE: 06/04/2021  
VIOLATION CODE: BUI  
ARREST TIME: 15:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:56

SUBJECT'S LAST NAME: ELVIR-FERMAN

DATE/TIME OF TEST: 08/31/2019 16:31:05

---

Diagnostics Check	OK	16:25
Air Blank	0.000	16:25
Control Test	0.081	16:25
Air Blank	0.000	16:26
Subject Sample # 1	0.089	16:26
Breath Volume	2.753	16:26
Air Blank	0.000	16:27
Air Blank	0.000	16:29
Subject Sample # 2	0.087	16:29
Breath Volume	2.679	16:29
Air Blank	0.000	16:30
Control Test	0.079	16:30
Air Blank	0.000	16:30
Diagnostics Check	OK	16:31

---

Data Download Date/Time: 09/11/2019 16:29

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO:	FWC	Date or Time Changed:	No
SOFTWARE VERSION:	8100.27	DRY GAS STANDARD LOT #:	962587
BREATH TEST OPERATOR:	CORBITT,KYLE,L	DRY GAS EXPIRY DATE:	06/04/2021
BREATH TEST OPERATOR AGENCY:	Okaloosa County SO	VIOLATION CODE:	BUI
ARRESTING OFFICER:	DONOVAN,CHARLES	ARREST TIME:	17:56
ARRESTING OFFICER AGENCY:	Okaloosa County SO	GENDER:	Female
LAST AGENCY INSPECTION:	08/22/2019	OBSERVATION PERIOD BEGAN:	18:24

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 08/31/2019 19:00:05

---

Diagnostics Check		OK	18:48	
Air Blank		0.000	18:48	
Control Test		0.081	18:49	
Air Blank		0.000	18:49	
Subject Sample	# 1	0.209	18:52	Volume Not Met
Breath Volume		0.671	18:52	
Air Blank		0.000	18:53	
Air Blank		0.000	18:55	
Subject Sample	# 2	0.218	18:58	Volume Not Met
Breath Volume		0.039	18:58	
Air Blank		0.000	18:59	
Control Test		0.083	18:59	
Air Blank		0.000	18:59	
Diagnostics Check		OK	19:00	

---

Data Download Date/Time: 09/11/2019 16:29



**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 962587
BREATH TEST OPERATOR: CORBITT,KYLE,L	DRY GAS EXPIRY DATE: 06/04/2021
BREATH TEST OPERATOR AGENCY: Okaloosa County SO	VIOLATION CODE: BUI
ARRESTING OFFICER: DONOVAN,CHARLES	ARREST TIME: 17:56
ARRESTING OFFICER AGENCY: Okaloosa County SO	GENDER: Female
LAST AGENCY INSPECTION: 08/22/2019	OBSERVATION PERIOD BEGAN: 18:24
SUBJECT'S LAST NAME: MOORE	DATE/TIME OF TEST: 08/31/2019 19:06:06

---

Diagnostics Check	OK	19:02	
Air Blank	0.000	19:02	
Control Test	0.081	19:02	
Air Blank	0.000	19:03	
Subject Sample	# 1	19:04	Subject Test Refused
Breath Volume	0.015	19:04	
Air Blank	0.000	19:05	
Control Test	0.081	19:05	
Air Blank	0.000	19:06	
Diagnostics Check	OK	19:06	

---

Data Download Date/Time: 09/11/2019 16:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MULLINS,JACKSON,K  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: CORBIN,BRITTON  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 08/22/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 962587  
DRY GAS EXPIRY DATE: 06/04/2021  
VIOLATION CODE: BUI  
ARREST TIME: 18:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:42

SUBJECT'S LAST NAME: BRUNO

DATE/TIME OF TEST: 08/31/2019 19:31:30

---

Diagnostics Check	OK	19:28	
Air Blank	0.000	19:28	
Control Test	0.081	19:28	
Air Blank	0.000	19:29	
Subject Sample	# 1	19:30	Subject Test Refused
Breath Volume	0.000	19:30	
Air Blank	0.000	19:30	
Control Test	0.081	19:30	
Air Blank	0.000	19:31	
Diagnostics Check	OK	19:31	

---

Data Download Date/Time: 09/11/2019 16:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000901

INSTRUMENT REGISTERED TO: FFWCC TAMPA  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PULASKI,DAMON,J  
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)  
ARRESTING OFFICER: PULASKI,DAMON  
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)  
LAST AGENCY INSPECTION: 07/26/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: EGGEN

DATE/TIME OF TEST: 08/13/2019 22:33:59

---

Diagnostics Check	OK	22:20	
Air Blank	0.000	22:21	
Control Test	0.078	22:21	
Air Blank	0.000	22:22	
Subject Sample	# 1	22:25	No Sample Provided
Breath Volume	0.000	22:25	
Air Blank	0.000	22:25	
Air Blank	0.000	22:27	
Subject Sample	# 2	0.142	22:28
Breath Volume	1.554	22:28	
Air Blank	0.000	22:29	
Air Blank	0.000	22:31	
Subject Sample	# 3	0.136	22:32
Breath Volume	1.992	22:32	
Air Blank	0.000	22:33	
Control Test	0.078	22:33	
Air Blank	0.000	22:33	
Diagnostics Check	OK	22:33	

---

Data Download Date/Time: 08/24/2019 13:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000903

INSTRUMENT REGISTERED TO: FFWCC SOUTH REGION  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRODBECK,JAMES,J  
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)  
ARRESTING OFFICER: BRODBECK,JAMES  
ARRESTING OFFICER AGENCY: FFWCC (West Palm)  
LAST AGENCY INSPECTION: 07/29/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 181732  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: BUI  
ARREST TIME: 17:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: BRITO GARCIA

DATE/TIME OF TEST: 08/18/2019 17:38:44

---

Diagnostics Check	OK	17:35	
Air Blank	0.000	17:36	
Control Test	0.077	17:36	
Air Blank	0.000	17:37	
Subject Sample	# 1	17:37	Subject Test Refused
Breath Volume	0.000	17:37	
Air Blank	0.000	17:37	
Control Test	0.077	17:38	
Air Blank	0.000	17:38	
Diagnostics Check	OK	17:38	

---

Data Download Date/Time: 08/29/2019 17:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000904

INSTRUMENT REGISTERED TO: FFWCC SOUTH REGION  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMPSON, WILLIAM, J  
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)  
ARRESTING OFFICER: ALVAREZ, MICHAEL  
ARRESTING OFFICER AGENCY: FFWCC (West Palm)  
LAST AGENCY INSPECTION: 07/31/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 666491  
DRY GAS EXPIRY DATE: 12/30/2019  
VIOLATION CODE: BUI  
ARREST TIME: 18:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:51

SUBJECT'S LAST NAME: CAREAGA

DATE/TIME OF TEST: 08/03/2019 20:02:36

---

Diagnostics Check	OK	19:56
Air Blank	0.000	19:56
Control Test	0.082	19:56
Air Blank	0.000	19:57
Subject Sample # 1	0.163	19:58
Breath Volume	2.964	19:58
Air Blank	0.000	19:58
Air Blank	0.000	20:00
Subject Sample # 2	0.153	20:01
Breath Volume	3.019	20:01
Air Blank	0.000	20:01
Control Test	0.081	20:02
Air Blank	0.000	20:02
Diagnostics Check	OK	20:02

---

Data Download Date/Time: 08/28/2019 01:00

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000906

INSTRUMENT REGISTERED TO: FISH & WILDLIFE CC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KLEIS,KEVIN,W  
BREATH TEST OPERATOR AGENCY: FFWCC (SOUTH A)  
ARRESTING OFFICER: YUREWITCH,KYLE  
ARRESTING OFFICER AGENCY: FFWCC (SOUTH A)  
LAST AGENCY INSPECTION: 08/01/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 666491  
DRY GAS EXPIRY DATE: 12/30/2019  
VIOLATION CODE: BUI  
ARREST TIME: 17:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:22

SUBJECT'S LAST NAME: MOULTON

DATE/TIME OF TEST: 08/17/2019 17:59:41

---

Diagnostics Check	OK	17:53
Air Blank	0.000	17:53
Control Test	0.078	17:54
Air Blank	0.000	17:54
Subject Sample # 1	0.085	17:55
Breath Volume	3.015	17:55
Air Blank	0.000	17:55
Air Blank	0.000	17:57
Subject Sample # 2	0.074	17:58
Breath Volume	1.273	17:58
Air Blank	0.000	17:58
Control Test	0.079	17:59
Air Blank	0.000	17:59
Diagnostics Check	OK	17:59

---

Data Download Date/Time: 09/03/2019 11:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SANTANA,ELIER,  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: SANTANA,ELIER  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 07/29/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 867139  
DRY GAS EXPIRY DATE: 11/15/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: CASTANEDA

DATE/TIME OF TEST: 08/03/2019 23:41:44

---

Diagnostics Check	OK	23:33
Air Blank	0.000	23:34
Control Test	0.080	23:34
Air Blank	0.000	23:34
Subject Sample # 1	0.211	23:36
Breath Volume	1.335	23:36
Air Blank	0.000	23:37
Air Blank	0.000	23:38
Subject Sample # 2	0.213	23:40
Breath Volume	1.191	23:40
Air Blank	0.000	23:40
Control Test	0.079	23:41
Air Blank	0.000	23:41
Diagnostics Check	OK	23:41

---

Data Download Date/Time: 08/15/2019 11:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: QUICK,BENJAMIN,  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: QUICK,BENJAMIN  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 07/29/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 867139  
DRY GAS EXPIRY DATE: 11/15/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: GUTIERREZ XILOJ

DATE/TIME OF TEST: 08/09/2019 23:47:25

---

Diagnostics Check	OK	23:40
Air Blank	0.000	23:41
Control Test	0.080	23:41
Air Blank	0.000	23:42
Subject Sample # 1	0.150	23:42
Breath Volume	3.015	23:42
Air Blank	0.000	23:43
Air Blank	0.000	23:45
Subject Sample # 2	0.133	23:45
Breath Volume	2.464	23:45
Air Blank	0.000	23:46
Control Test	0.080	23:46
Air Blank	0.000	23:47
Diagnostics Check	OK	23:47

---

Data Download Date/Time: 08/15/2019 11:12



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: MONTEMAGNO,CARLO  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 07/29/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 867139  
DRY GAS EXPIRY DATE: 11/15/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:47  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: BARTON

DATE/TIME OF TEST: 08/16/2019 00:27:25

---

Diagnostics Check	OK	00:24	
Air Blank	0.000	00:25	
Control Test	0.080	00:25	
Air Blank	0.000	00:25	
Subject Sample	# 1	00:26	Subject Test Refused
Breath Volume	0.000	00:26	
Air Blank	0.000	00:26	
Control Test	0.080	00:26	
Air Blank	0.000	00:27	
Diagnostics Check	OK	00:27	

---

Data Download Date/Time: 08/26/2019 15:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KING,GRACE,N  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: MONTEMAGNO,CARLO  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 07/29/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 867139  
DRY GAS EXPIRY DATE: 11/15/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: SHEROOD

DATE/TIME OF TEST: 08/21/2019 03:13:16

---

Diagnostics Check	OK	03:06
Air Blank	0.000	03:07
Control Test	0.081	03:07
Air Blank	0.000	03:08
Subject Sample # 1	0.000	03:08
Breath Volume	1.265	03:08
Air Blank	0.000	03:09
Air Blank	0.000	03:11
Subject Sample # 2	0.000	03:11
Breath Volume	1.500	03:11
Air Blank	0.000	03:12
Control Test	0.081	03:12
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

---

Data Download Date/Time: 08/26/2019 15:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: MONTEMAGNO,CARLO  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 07/29/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 867139  
DRY GAS EXPIRY DATE: 11/15/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: HERNANDEZ MENDOZA

DATE/TIME OF TEST: 08/23/2019 23:02:50

---

Diagnostics Check	OK	23:00	
Air Blank	0.000	23:00	
Control Test	0.080	23:00	
Air Blank	0.000	23:01	
Subject Sample	# 1	23:01	Subject Test Refused
Breath Volume	0.000	23:01	
Air Blank	0.000	23:01	
Control Test	0.079	23:02	
Air Blank	0.000	23:02	
Diagnostics Check	OK	23:02	

---

Data Download Date/Time: 08/26/2019 15:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: MONTEMAGNO,CARLO  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 07/29/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 867139  
DRY GAS EXPIRY DATE: 11/15/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:16

SUBJECT'S LAST NAME: GOMEZOSUMA

DATE/TIME OF TEST: 08/24/2019 22:44:35

---

Diagnostics Check	OK	22:37
Air Blank	0.000	22:38
Control Test	0.080	22:38
Air Blank	0.000	22:38
Subject Sample # 1	0.225	22:39
Breath Volume	2.238	22:39
Air Blank	0.000	22:40
Air Blank	0.000	22:41
Subject Sample # 2	0.222	22:43
Breath Volume	2.683	22:43
Air Blank	0.000	22:43
Control Test	0.079	22:44
Air Blank	0.000	22:44
Diagnostics Check	OK	22:44

---

Data Download Date/Time: 08/26/2019 15:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: MONTEMAGNO,CARLO  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 07/29/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 867139  
DRY GAS EXPIRY DATE: 11/15/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: SQUIER

DATE/TIME OF TEST: 08/25/2019 03:51:22

---

Diagnostics Check	OK	03:48	
Air Blank	0.000	03:48	
Control Test	0.080	03:49	
Air Blank	0.000	03:49	
Subject Sample	# 1	03:50	Subject Test Refused
Breath Volume	0.000	03:50	
Air Blank	0.000	03:50	
Control Test	0.081	03:50	
Air Blank	0.000	03:51	
Diagnostics Check	OK	03:51	

---

Data Download Date/Time: 08/26/2019 15:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: QUICK,BENJAMIN,  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: QUICK,BENJAMIN  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 07/29/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 867139  
DRY GAS EXPIRY DATE: 11/15/2020  
VIOLATION CODE: DUI  
ARREST TIME: 18:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:44

SUBJECT'S LAST NAME: MENDOZA

DATE/TIME OF TEST: 08/25/2019 19:11:02

---

Diagnostics Check	OK	19:04
Air Blank	0.000	19:05
Control Test	0.080	19:05
Air Blank	0.000	19:06
Subject Sample # 1	0.224	19:06
Breath Volume	1.199	19:06
Air Blank	0.000	19:07
Air Blank	0.000	19:08
Subject Sample # 2	0.206	19:09
Breath Volume	1.164	19:09
Air Blank	0.000	19:10
Control Test	0.081	19:10
Air Blank	0.000	19:10
Diagnostics Check	OK	19:11

---

Data Download Date/Time: 08/26/2019 15:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: LEDFORD,TAYLOR  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 08/26/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 867139  
DRY GAS EXPIRY DATE: 11/15/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: RAPALO ORELLANA

DATE/TIME OF TEST: 08/29/2019 02:36:01

---

Diagnostics Check	OK	02:29
Air Blank	0.000	02:29
Control Test	0.079	02:30
Air Blank	0.000	02:30
Subject Sample # 1	0.147	02:31
Breath Volume	3.203	02:31
Air Blank	0.000	02:31
Air Blank	0.000	02:33
Subject Sample # 2	0.151	02:34
Breath Volume	3.257	02:34
Air Blank	0.000	02:35
Control Test	0.079	02:35
Air Blank	0.000	02:35
Diagnostics Check	OK	02:36

---

Data Download Date/Time: 09/17/2019 10:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ACKERMAN,BRIAN,K  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: ACKERMAN,BRIAN  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 08/26/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 867139  
DRY GAS EXPIRY DATE: 11/15/2020  
VIOLATION CODE: DUI  
ARREST TIME: 18:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: EDDY

DATE/TIME OF TEST: 08/31/2019 19:53:57

---

Diagnostics Check	OK	19:51	
Air Blank	0.000	19:51	
Control Test	0.079	19:51	
Air Blank	0.000	19:52	
Subject Sample	# 1	19:52	Subject Test Refused
Breath Volume	0.000	19:52	
Air Blank	0.000	19:53	
Control Test	0.079	19:53	
Air Blank	0.000	19:53	
Diagnostics Check	OK	19:53	

---

Data Download Date/Time: 09/17/2019 10:18



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000908

INSTRUMENT REGISTERED TO: PORT ST JOE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GERBER,OLIVER,B  
BREATH TEST OPERATOR AGENCY: FFWCC (NW REGION)  
ARRESTING OFFICER: GERBER,OLIVER  
ARRESTING OFFICER AGENCY: FFWCC (NW REGION)  
LAST AGENCY INSPECTION: 08/19/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08819080A1  
DRY GAS EXPIRY DATE: 06/05/2021  
VIOLATION CODE: BUI  
ARREST TIME: 17:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: MIDDLEMAS

DATE/TIME OF TEST: 08/24/2019 18:43:01

---

Diagnostics Check	OK	18:39	
Air Blank	0.000	18:40	
Control Test	0.079	18:40	
Air Blank	0.000	18:41	
Subject Sample	# 1	18:41	Subject Test Refused
Breath Volume	0.000	18:41	
Air Blank	0.000	18:42	
Control Test	0.079	18:42	
Air Blank	0.000	18:42	
Diagnostics Check	OK	18:43	

---

Data Download Date/Time: 09/16/2019 14:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TISDALE,KATIE,E  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: KUPKA,RIKKI  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 07/19/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: SERRANO PENA

DATE/TIME OF TEST: 08/01/2019 22:28:58

---

Diagnostics Check	OK	22:22
Air Blank	0.000	22:22
Control Test	0.079	22:23
Air Blank	0.000	22:23
Subject Sample # 1	0.157	22:24
Breath Volume	2.656	22:24
Air Blank	0.000	22:25
Air Blank	0.000	22:26
Subject Sample # 2	0.161	22:27
Breath Volume	3.722	22:27
Air Blank	0.000	22:28
Control Test	0.079	22:28
Air Blank	0.000	22:28
Diagnostics Check	OK	22:28

---

Data Download Date/Time: 08/14/2019 12:02

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BROWN,EARTHEN,A  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: BROWN,EARTHEN  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 07/19/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: CHERISOL

DATE/TIME OF TEST: 08/03/2019 04:32:26

---

Diagnostics Check	OK	04:23
Air Blank	0.000	04:23
Control Test	0.080	04:24
Air Blank	0.000	04:24
Subject Sample # 1	0.198	04:25
Breath Volume	1.425	04:25
Air Blank	0.000	04:26
Air Blank	0.000	04:28
Subject Sample # 2	0.196	04:30
Breath Volume	1.441	04:30
Air Blank	0.000	04:31
Control Test	0.080	04:31
Air Blank	0.000	04:32
Diagnostics Check	OK	04:32

---

Data Download Date/Time: 08/14/2019 12:02

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HERON,LAUREN,A  
BREATH TEST OPERATOR AGENCY: Sanibel PD  
ARRESTING OFFICER: HERON,L  
ARRESTING OFFICER AGENCY: Sanibel PD  
LAST AGENCY INSPECTION: 07/19/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: GATES

DATE/TIME OF TEST: 08/04/2019 01:36:07

---

Diagnostics Check	OK	01:29
Air Blank	0.000	01:29
Control Test	0.079	01:30
Air Blank	0.000	01:30
Subject Sample # 1	0.189	01:31
Breath Volume	2.500	01:31
Air Blank	0.000	01:32
Air Blank	0.000	01:34
Subject Sample # 2	0.185	01:34
Breath Volume	2.628	01:34
Air Blank	0.000	01:35
Control Test	0.076	01:35
Air Blank	0.000	01:36
Diagnostics Check	OK	01:36

---

Data Download Date/Time: 08/14/2019 12:02

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CANFIELD,A.,B  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: MCGILL,K.  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 07/19/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 08/09/2019 00:05:15

---

Diagnostics Check	OK	00:02	
Air Blank	0.000	00:02	
Control Test	0.079	00:03	
Air Blank	0.000	00:03	
Subject Sample	# 1	00:03	Subject Test Refused
Breath Volume	0.000	00:03	
Air Blank	0.000	00:04	
Control Test	0.078	00:04	
Air Blank	0.000	00:05	
Diagnostics Check	OK	00:05	

---

Data Download Date/Time: 08/14/2019 12:02

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CANFIELD 18096,A.,B  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: MCGILL,K  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 07/19/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 08/09/2019 00:35:51

---

Diagnostics Check		OK	00:28
Air Blank		0.000	00:28
Control Test		0.079	00:29
Air Blank		0.000	00:29
Subject Sample	# 1	0.264	00:30
Breath Volume		2.363	00:30
Air Blank		0.000	00:31
Air Blank		0.000	00:33
Subject Sample	# 2	0.279	00:34
Breath Volume		2.195	00:34
Air Blank		0.000	00:34
Control Test		0.077	00:35
Air Blank		0.000	00:35
Diagnostics Check		OK	00:35

---

Data Download Date/Time: 08/14/2019 12:02

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WINCELOWICZ,LARRY,R  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: HOYTE,J  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 07/19/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: DULIN

DATE/TIME OF TEST: 08/09/2019 23:38:25

---

Diagnostics Check	OK	23:32
Air Blank	0.000	23:32
Control Test	0.080	23:33
Air Blank	0.000	23:33
Subject Sample # 1	0.132	23:34
Breath Volume	1.789	23:34
Air Blank	0.000	23:34
Air Blank	0.000	23:36
Subject Sample # 2	0.134	23:36
Breath Volume	1.765	23:36
Air Blank	0.000	23:37
Control Test	0.079	23:37
Air Blank	0.000	23:38
Diagnostics Check	OK	23:38

---

Data Download Date/Time: 08/14/2019 12:02

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTINS,KYLE,L  
BREATH TEST OPERATOR AGENCY: Fort Myers PD  
ARRESTING OFFICER: WERBALIS,JOANN  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 07/19/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: PALUMBO

DATE/TIME OF TEST: 08/11/2019 00:49:24

---

Diagnostics Check	OK	00:43
Air Blank	0.000	00:43
Control Test	0.080	00:43
Air Blank	0.000	00:44
Subject Sample # 1	0.000	00:45
Breath Volume	4.460	00:45
Air Blank	0.000	00:45
Air Blank	0.000	00:47
Subject Sample # 2	0.000	00:48
Breath Volume	4.160	00:48
Air Blank	0.000	00:48
Control Test	0.080	00:48
Air Blank	0.000	00:49
Diagnostics Check	OK	00:49

---

Data Download Date/Time: 08/14/2019 12:02



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTINS,KYLE,L  
BREATH TEST OPERATOR AGENCY: Fort Myers PD  
ARRESTING OFFICER: VASALLO,JASMINE  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: NORMAN

DATE/TIME OF TEST: 08/14/2019 23:59:36

---

Diagnostics Check	OK	23:53
Air Blank	0.000	23:53
Control Test	0.080	23:54
Air Blank	0.000	23:54
Subject Sample # 1	0.000	23:55
Breath Volume	3.332	23:55
Air Blank	0.000	23:55
Air Blank	0.000	23:57
Subject Sample # 2	0.000	23:58
Breath Volume	3.722	23:58
Air Blank	0.000	23:58
Control Test	0.079	23:59
Air Blank	0.000	23:59
Diagnostics Check	OK	23:59

---

Data Download Date/Time: 09/18/2019 12:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTINS,KYLE,L  
BREATH TEST OPERATOR AGENCY: Fort Myers PD  
ARRESTING OFFICER: MORGAN,ROHAN  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 08/15/2019 03:10:25

---

Diagnostics Check		OK	03:03
Air Blank		0.000	03:04
Control Test		0.079	03:04
Air Blank		0.000	03:05
Subject Sample	# 1	0.154	03:06
Breath Volume		2.601	03:06
Air Blank		0.000	03:06
Air Blank		0.000	03:08
Subject Sample	# 2	0.159	03:08
Breath Volume		2.570	03:08
Air Blank		0.000	03:09
Control Test		0.079	03:09
Air Blank		0.000	03:10
Diagnostics Check		OK	03:10

---

Data Download Date/Time: 09/18/2019 12:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WAY,J,T  
BREATH TEST OPERATOR AGENCY: Fort Myers PD  
ARRESTING OFFICER: KUHLM,J  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 15:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:58

SUBJECT'S LAST NAME: SAXTON

DATE/TIME OF TEST: 08/16/2019 16:27:03

---

Diagnostics Check	OK	16:20
Air Blank	0.000	16:21
Control Test	0.079	16:21
Air Blank	0.000	16:22
Subject Sample # 1	0.210	16:22
Breath Volume	2.214	16:22
Air Blank	0.000	16:23
Air Blank	0.000	16:25
Subject Sample # 2	0.198	16:25
Breath Volume	2.750	16:25
Air Blank	0.000	16:26
Control Test	0.079	16:26
Air Blank	0.000	16:26
Diagnostics Check	OK	16:27

---

Data Download Date/Time: 09/18/2019 12:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CANFIELD 18096,A.,B  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: PAHLKE,M  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: AUGUSTINE

DATE/TIME OF TEST: 08/18/2019 02:18:14

---

Diagnostics Check	OK	02:11
Air Blank	0.000	02:11
Control Test	0.079	02:12
Air Blank	0.000	02:12
Subject Sample # 1	0.147	02:13
Breath Volume	2.382	02:13
Air Blank	0.000	02:14
Air Blank	0.000	02:16
Subject Sample # 2	0.150	02:16
Breath Volume	1.937	02:16
Air Blank	0.000	02:17
Control Test	0.079	02:17
Air Blank	0.000	02:18
Diagnostics Check	OK	02:18

---

Data Download Date/Time: 09/18/2019 12:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WAY,J,T  
BREATH TEST OPERATOR AGENCY: Fort Myers PD  
ARRESTING OFFICER: WAY,J  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 10:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 12:28

SUBJECT'S LAST NAME: SINGLETARY

DATE/TIME OF TEST: 08/18/2019 12:54:38

---

Diagnostics Check	OK	12:48
Air Blank	0.000	12:48
Control Test	0.080	12:49
Air Blank	0.000	12:49
Subject Sample # 1	0.040	12:50
Breath Volume	3.093	12:50
Air Blank	0.000	12:50
Air Blank	0.000	12:52
Subject Sample # 2	0.038	12:53
Breath Volume	3.414	12:53
Air Blank	0.000	12:53
Control Test	0.079	12:54
Air Blank	0.000	12:54
Diagnostics Check	OK	12:54

---

Data Download Date/Time: 09/18/2019 12:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMPSON,SCOTT,A  
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)  
ARRESTING OFFICER: THOMPSON,SCOTT  
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)  
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 08/20/2019 00:04:41

---

Diagnostics Check	OK	00:00	
Air Blank	0.000	00:01	
Control Test	0.079	00:01	
Air Blank	0.000	00:02	
Subject Sample	# 1	00:03	RFI Detect
Breath Volume	0.644	00:03	
Air Blank		00:04	RFI Detect
Air Blank	0.000	00:04	

---

Data Download Date/Time: 09/18/2019 12:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMPSON,SCOTT,A  
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)  
ARRESTING OFFICER: THOMPSON,SCOTT  
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)  
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 08/20/2019 00:13:18

---

Diagnostics Check	OK	00:06
Air Blank	0.000	00:07
Control Test	0.078	00:07
Air Blank	0.000	00:07
Subject Sample # 1	0.174	00:08
Breath Volume	1.937	00:08
Air Blank	0.000	00:09
Air Blank	0.000	00:11
Subject Sample # 2	0.168	00:11
Breath Volume	1.402	00:11
Air Blank	0.000	00:12
Control Test	0.078	00:12
Air Blank	0.000	00:13
Diagnostics Check	OK	00:13

---

Data Download Date/Time: 09/18/2019 12:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOULD,LEONARD,  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: GOULD,LEONARD  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 17:03  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 17:55

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 08/21/2019 18:27:08

---

Diagnostics Check	OK	18:20
Air Blank	0.000	18:21
Control Test	0.079	18:21
Air Blank	0.000	18:22
Subject Sample # 1	0.071	18:22
Breath Volume	1.757	18:22
Air Blank	0.000	18:23
Air Blank	0.000	18:25
Subject Sample # 2	0.068	18:25
Breath Volume	1.269	18:25
Air Blank	0.000	18:26
Control Test	0.079	18:26
Air Blank	0.000	18:27
Diagnostics Check	OK	18:27

---

Data Download Date/Time: 09/18/2019 12:37



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CANFIELD18096,A.,B  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: LEON,B  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 16:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:27

SUBJECT'S LAST NAME: BARTOS

DATE/TIME OF TEST: 08/22/2019 18:06:02

---

Diagnostics Check	OK	18:00
Air Blank	0.000	18:00
Control Test	0.080	18:00
Air Blank	0.000	18:01
Subject Sample # 1	0.000	18:01
Breath Volume	2.539	18:01
Air Blank	0.000	18:02
Air Blank	0.000	18:04
Subject Sample # 2	0.000	18:04
Breath Volume	2.093	18:04
Air Blank	0.000	18:05
Control Test	0.079	18:05
Air Blank	0.000	18:05
Diagnostics Check	OK	18:06

---

Data Download Date/Time: 09/18/2019 12:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TISDALE,KATIE,E  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: LOUIS,RENET  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: AYLOR

DATE/TIME OF TEST: 08/23/2019 22:40:53

---

Diagnostics Check	OK	22:34
Air Blank	0.000	22:35
Control Test	0.080	22:35
Air Blank	0.000	22:35
Subject Sample # 1	0.219	22:36
Breath Volume	3.363	22:36
Air Blank	0.000	22:37
Air Blank	0.000	22:38
Subject Sample # 2	0.209	22:39
Breath Volume	3.039	22:39
Air Blank	0.000	22:39
Control Test	0.079	22:40
Air Blank	0.000	22:40
Diagnostics Check	OK	22:40

---

Data Download Date/Time: 09/18/2019 12:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WINCELOWICZ,LARRY,R  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: WINCELOWICZ,LARRY  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: DE LEON POOU

DATE/TIME OF TEST: 08/24/2019 00:03:16

---

Diagnostics Check	OK	23:56
Air Blank	0.000	23:56
Control Test	0.079	23:57
Air Blank	0.000	23:57
Subject Sample # 2	0.141	23:58
Breath Volume	1.710	23:58
Air Blank	0.000	23:59
Air Blank	0.000	00:01
Subject Sample # 1	0.132	00:01
Breath Volume	1.203	00:01
Air Blank	0.000	00:02
Control Test	0.079	00:02
Air Blank	0.000	00:03
Diagnostics Check	OK	00:03

---

Data Download Date/Time: 09/18/2019 12:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TISDALE,KATIE,E  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: BREDECK,JORDAN  
ARRESTING OFFICER AGENCY: Lee Port Authority  
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:27  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: MIGLIORESE COLLARD

DATE/TIME OF TEST: 08/24/2019 02:05:37

---

Diagnostics Check	OK	02:02	
Air Blank	0.000	02:02	
Control Test	0.079	02:03	
Air Blank	0.000	02:03	
Subject Sample	# 1	02:04	Slope Not Met
Breath Volume	1.105	02:04	
Air Blank	0.000	02:04	
Control Test	0.079	02:05	
Air Blank	0.000	02:05	
Diagnostics Check	OK	02:05	

---

Data Download Date/Time: 09/18/2019 12:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TISDALE,KATIE,E  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: LEON,BRIAN  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: CASTANEDA

DATE/TIME OF TEST: 08/24/2019 04:45:22

---

Diagnostics Check	OK	04:38
Air Blank	0.000	04:38
Control Test	0.079	04:39
Air Blank	0.000	04:39
Subject Sample # 1	0.070	04:40
Breath Volume	2.050	04:40
Air Blank	0.000	04:41
Air Blank	0.000	04:43
Subject Sample # 2	0.070	04:43
Breath Volume	3.128	04:43
Air Blank	0.000	04:44
Control Test	0.079	04:44
Air Blank	0.000	04:45
Diagnostics Check	OK	04:45

---

Data Download Date/Time: 09/18/2019 12:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WINCELOWICZ,LARRY,R  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: WINCELOWICZ,LARRY  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: MCDONALD

DATE/TIME OF TEST: 08/25/2019 00:41:20

---

Diagnostics Check	OK	00:34
Air Blank	0.000	00:35
Control Test	0.080	00:35
Air Blank	0.000	00:35
Subject Sample # 1	0.059	00:36
Breath Volume	1.417	00:36
Air Blank	0.000	00:36
Air Blank	0.000	00:38
Subject Sample # 2	0.060	00:39
Breath Volume	1.835	00:39
Air Blank	0.000	00:40
Control Test	0.080	00:40
Air Blank	0.000	00:41
Diagnostics Check	OK	00:41

---

Data Download Date/Time: 09/18/2019 12:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROMAN,A,C  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: SEMIKLOSE,C  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: LAGOS ESPINOZA

DATE/TIME OF TEST: 08/25/2019 05:11:37

---

Diagnostics Check	OK	05:05
Air Blank	0.000	05:05
Control Test	0.080	05:06
Air Blank	0.000	05:06
Subject Sample # 1	0.207	05:07
Breath Volume	1.339	05:07
Air Blank	0.000	05:07
Air Blank	0.000	05:09
Subject Sample # 2	0.203	05:10
Breath Volume	2.417	05:10
Air Blank	0.000	05:10
Control Test	0.079	05:11
Air Blank	0.000	05:11
Diagnostics Check	OK	05:11

---

Data Download Date/Time: 09/18/2019 12:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CANFIELD 18096,A.,B  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: THOMPSON,R.  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:35  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: SEWELL

DATE/TIME OF TEST: 08/27/2019 00:55:31

---

Diagnostics Check		OK	00:49
Air Blank		0.000	00:49
Control Test		0.080	00:50
Air Blank		0.000	00:50
Subject Sample	# 1	0.184	00:51
Breath Volume		2.386	00:51
Air Blank		0.000	00:51
Air Blank		0.000	00:53
Subject Sample	# 2	0.187	00:54
Breath Volume		2.472	00:54
Air Blank		0.000	00:54
Control Test		0.079	00:54
Air Blank		0.000	00:55
Diagnostics Check		OK	00:55

---

Data Download Date/Time: 09/18/2019 12:37



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CANFIELD 18096,A.,B  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: KING,J  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: COMBS

DATE/TIME OF TEST: 08/27/2019 04:13:41

---

Diagnostics Check	OK	04:06
Air Blank	0.000	04:07
Control Test	0.079	04:07
Air Blank	0.000	04:08
Subject Sample # 1	0.048	04:08
Breath Volume	1.898	04:08
Air Blank	0.000	04:09
Air Blank	0.000	04:11
Subject Sample # 2	0.049	04:12
Breath Volume	2.765	04:12
Air Blank	0.000	04:12
Control Test	0.080	04:13
Air Blank	0.000	04:13
Diagnostics Check	OK	04:13

---

Data Download Date/Time: 09/18/2019 12:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TISDALE,KATIE,E  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: SELIMOVIC,ADNAN  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: HERNANDEZ-SUBIT

DATE/TIME OF TEST: 08/28/2019 20:39:46

---

Diagnostics Check	OK	20:27	
Air Blank	0.000	20:28	
Control Test	0.079	20:28	
Air Blank	0.000	20:29	
Subject Sample # 1	0.337	20:29	
Breath Volume	3.312	20:29	
Air Blank	0.000	20:30	
Air Blank	0.000	20:32	
Subject Sample # 2		20:35	No Sample Provided
Breath Volume	0.000	20:35	
Air Blank	0.000	20:35	
Air Blank	0.000	20:37	
Subject Sample # 3	0.349	20:38	
Breath Volume	1.601	20:38	
Air Blank	0.000	20:38	
Control Test	0.079	20:39	
Air Blank	0.000	20:39	
Diagnostics Check	OK	20:39	

---

Data Download Date/Time: 09/18/2019 12:38

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH,JOSIAH,R  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: SMITH,JOSIAH  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:23  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: KEISLING

DATE/TIME OF TEST: 08/29/2019 04:11:37

---

Diagnostics Check	OK	04:04
Air Blank	0.000	04:05
Control Test	0.079	04:05
Air Blank	0.000	04:06
Subject Sample # 1	0.178	04:06
Breath Volume	1.675	04:06
Air Blank	0.000	04:07
Air Blank	0.000	04:09
Subject Sample # 2	0.170	04:10
Breath Volume	1.156	04:10
Air Blank	0.000	04:10
Control Test	0.080	04:11
Air Blank	0.000	04:11
Diagnostics Check	OK	04:11

---

Data Download Date/Time: 09/18/2019 12:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTINS,KYLE,L  
BREATH TEST OPERATOR AGENCY: Fort Myers PD  
ARRESTING OFFICER: WERBALIS,JOANN  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: BARLOW

DATE/TIME OF TEST: 08/30/2019 00:59:12

---

Diagnostics Check		OK	00:49
Air Blank		0.000	00:50
Control Test		0.079	00:50
Air Blank		0.000	00:50
Subject Sample	# 1	0.189	00:51
Breath Volume		1.933	00:51
Air Blank		0.000	00:52
Air Blank		0.000	00:53
Subject Sample	# 2	0.218	00:54
Breath Volume		2.457	00:54
Air Blank		0.000	00:55
Air Blank		0.000	00:56
Subject Sample	# 3	0.238	00:57
Breath Volume		2.437	00:57
Air Blank		0.000	00:58
Control Test		0.078	00:58
Air Blank		0.000	00:59
Diagnostics Check		OK	00:59

---

Data Download Date/Time: 09/18/2019 12:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAWLINS,DUSTIN,P  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: CAIL,MATTHEW  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 09:37  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 10:35

SUBJECT'S LAST NAME: ANDOLINO

DATE/TIME OF TEST: 08/31/2019 11:13:40

---

Diagnostics Check	OK	11:10	
Air Blank	0.000	11:11	
Control Test	0.080	11:11	
Air Blank	0.000	11:11	
Subject Sample	# 1	11:12	Subject Test Refused
Breath Volume	0.000	11:12	
Air Blank	0.000	11:12	
Control Test	0.080	11:13	
Air Blank	0.000	11:13	
Diagnostics Check	OK	11:13	

---

Data Download Date/Time: 09/18/2019 12:38

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOULD,LEONARD,  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: GOULD,LEONARD  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 07/19/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: CASTILLO CLARO

DATE/TIME OF TEST: 08/03/2019 22:52:39

---

Diagnostics Check		OK	22:46
Air Blank		0.000	22:46
Control Test		0.080	22:46
Air Blank		0.000	22:47
Subject Sample	# 1	0.186	22:48
Breath Volume		2.000	22:48
Air Blank		0.000	22:48
Air Blank		0.000	22:50
Subject Sample	# 2	0.167	22:51
Breath Volume		1.386	22:51
Air Blank		0.000	22:51
Control Test		0.077	22:52
Air Blank		0.000	22:52
Diagnostics Check		OK	22:52

---

Data Download Date/Time: 08/14/2019 09:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HERON,LAUREN,A  
BREATH TEST OPERATOR AGENCY: Sanibel PD  
ARRESTING OFFICER: HERON,LAUREN  
ARRESTING OFFICER AGENCY: Sanibel PD  
LAST AGENCY INSPECTION: 07/19/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: GATES

DATE/TIME OF TEST: 08/04/2019 01:25:45

---

Diagnostics Check	OK	01:22	
Air Blank	0.000	01:23	
Control Test	0.080	01:23	
Air Blank	0.000	01:23	
Subject Sample	# 1 0.188	01:24	
Breath Volume	1.425	01:24	
Air Blank		01:25	Purge Fail
Air Blank	0.000	01:25	

---

Data Download Date/Time: 08/14/2019 09:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHILKE,CARL,E  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: SCHILKE,CARL  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 07/19/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 17:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: RAYMUNDO TERRAZA

DATE/TIME OF TEST: 08/04/2019 18:14:42

---

Diagnostics Check	OK	18:08
Air Blank	0.000	18:08
Control Test	0.079	18:09
Air Blank	0.000	18:09
Subject Sample # 1	0.157	18:10
Breath Volume	1.558	18:10
Air Blank	0.000	18:10
Air Blank	0.000	18:12
Subject Sample # 2	0.158	18:13
Breath Volume	2.894	18:13
Air Blank	0.000	18:13
Control Test	0.079	18:14
Air Blank	0.000	18:14
Diagnostics Check	OK	18:14

---

Data Download Date/Time: 08/14/2019 09:19



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCGOWAN,SAVANNAH,K  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: MCGOWAN,SAVANNAH  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 07/19/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 15:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: CARROLL

DATE/TIME OF TEST: 08/07/2019 17:19:05

---

Diagnostics Check	OK	17:12
Air Blank	0.000	17:12
Control Test	0.080	17:12
Air Blank	0.000	17:13
Subject Sample # 1	0.196	17:14
Breath Volume	1.925	17:14
Air Blank	0.000	17:15
Air Blank	0.000	17:16
Subject Sample # 2	0.203	17:17
Breath Volume	1.968	17:17
Air Blank	0.000	17:18
Control Test	0.078	17:18
Air Blank	0.000	17:19
Diagnostics Check	OK	17:19

---

Data Download Date/Time: 08/14/2019 09:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WITT,KELLY,  
BREATH TEST OPERATOR AGENCY: Fort Myers PD  
ARRESTING OFFICER: BRADFORD,HARLEE  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 07/19/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 08/07/2019 23:15:31

---

Diagnostics Check	OK	23:07
Air Blank	0.000	23:07
Control Test	0.081	23:08
Air Blank	0.000	23:08
Subject Sample # 1	0.048	23:11
Breath Volume	1.472	23:11
Air Blank	0.000	23:11
Air Blank	0.000	23:13
Subject Sample # 2	0.044	23:13
Breath Volume	2.570	23:13
Air Blank	0.000	23:14
Control Test	0.079	23:14
Air Blank	0.000	23:15
Diagnostics Check	OK	23:15

---

Data Download Date/Time: 08/14/2019 09:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHILKE,CARL,E  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: SCHILKE,CARL  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 07/19/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:18

SUBJECT'S LAST NAME: ELKINS

DATE/TIME OF TEST: 08/08/2019 20:44:52

---

Diagnostics Check	OK	20:38
Air Blank	0.000	20:38
Control Test	0.080	20:39
Air Blank	0.000	20:39
Subject Sample # 1	0.277	20:40
Breath Volume	2.308	20:40
Air Blank	0.000	20:41
Air Blank	0.000	20:42
Subject Sample # 2	0.277	20:43
Breath Volume	2.589	20:43
Air Blank	0.000	20:43
Control Test	0.080	20:44
Air Blank	0.000	20:44
Diagnostics Check	OK	20:44

---

Data Download Date/Time: 08/14/2019 09:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHILKE,CARL,E  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: BRAY,CADE  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 07/19/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: LIGAS

DATE/TIME OF TEST: 08/08/2019 21:38:32

---

Diagnostics Check	OK	21:32
Air Blank	0.000	21:32
Control Test	0.080	21:32
Air Blank	0.000	21:33
Subject Sample # 1	0.000	21:34
Breath Volume	4.621	21:34
Air Blank	0.000	21:34
Air Blank	0.000	21:36
Subject Sample # 2	0.000	21:37
Breath Volume	4.460	21:37
Air Blank	0.000	21:37
Control Test	0.080	21:37
Air Blank	0.000	21:38
Diagnostics Check	OK	21:38

---

Data Download Date/Time: 08/14/2019 09:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CANFIELD,A.,B  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: MCGILL,K.  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 07/19/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 08/08/2019 23:54:19

---

Diagnostics Check	OK	23:51	
Air Blank	0.000	23:51	
Control Test	0.080	23:51	
Air Blank	0.000	23:52	
Subject Sample	# 1 0.323	23:52	
Breath Volume	1.531	23:52	
Air Blank		23:53	Purge Fail
Air Blank	0.000	23:54	

---

Data Download Date/Time: 08/14/2019 09:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WITT,KELLY,  
BREATH TEST OPERATOR AGENCY: Fort Myers PD  
ARRESTING OFFICER: WITT,KELLY  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 07/19/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: GOHL

DATE/TIME OF TEST: 08/10/2019 02:06:01

---

Diagnostics Check	OK	01:58
Air Blank	0.000	01:58
Control Test	0.080	01:59
Air Blank	0.000	01:59
Subject Sample # 1	0.159	02:01
Breath Volume	1.808	02:01
Air Blank	0.000	02:02
Air Blank	0.000	02:03
Subject Sample # 2	0.153	02:04
Breath Volume	2.394	02:04
Air Blank	0.000	02:05
Control Test	0.078	02:05
Air Blank	0.000	02:05
Diagnostics Check	OK	02:06

---

Data Download Date/Time: 08/14/2019 09:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH, JOSIAH, R  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: SMITH, JOSIAH  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 07/19/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:55  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: OLMO

DATE/TIME OF TEST: 08/11/2019 04:54:48

---

Diagnostics Check	OK	04:48
Air Blank	0.000	04:48
Control Test	0.080	04:48
Air Blank	0.000	04:49
Subject Sample # 1	0.159	04:50
Breath Volume	2.519	04:50
Air Blank	0.000	04:50
Air Blank	0.000	04:52
Subject Sample # 2	0.159	04:53
Breath Volume	2.277	04:53
Air Blank	0.000	04:53
Control Test	0.079	04:54
Air Blank	0.000	04:54
Diagnostics Check	OK	04:54

---

Data Download Date/Time: 08/14/2019 09:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WAY,JOSEPH,T  
BREATH TEST OPERATOR AGENCY: Fort Myers PD  
ARRESTING OFFICER: WAY,JOSEPH  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 07/19/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 09:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:44

SUBJECT'S LAST NAME: LANDINO

DATE/TIME OF TEST: 08/12/2019 10:11:06

---

Diagnostics Check	OK	10:04
Air Blank	0.000	10:04
Control Test	0.080	10:05
Air Blank	0.000	10:05
Subject Sample # 1	0.000	10:06
Breath Volume	2.304	10:06
Air Blank	0.000	10:07
Air Blank	0.000	10:09
Subject Sample # 2	0.000	10:09
Breath Volume	1.480	10:09
Air Blank	0.000	10:10
Control Test	0.079	10:10
Air Blank	0.000	10:11
Diagnostics Check	OK	10:11

---

Data Download Date/Time: 08/14/2019 09:19



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KONIECZKI,D,  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: WALSH,L  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 07/19/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 07:04  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 07:45

SUBJECT'S LAST NAME: LUCKEY

DATE/TIME OF TEST: 08/13/2019 08:17:46

---

Diagnostics Check		OK	08:10
Air Blank		0.000	08:11
Control Test		0.080	08:11
Air Blank		0.000	08:12
Subject Sample	# 1	0.000	08:13
Breath Volume		2.265	08:13
Air Blank		0.000	08:13
Air Blank		0.000	08:15
Subject Sample	# 2	0.000	08:16
Breath Volume		2.175	08:16
Air Blank		0.000	08:16
Control Test		0.080	08:17
Air Blank		0.000	08:17
Diagnostics Check		OK	08:17

---

Data Download Date/Time: 08/14/2019 09:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LINDHOLM,JOHN,T  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: LINDHOLM,JOHN  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 07/19/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 08/14/2019 02:11:06

---

Diagnostics Check	OK	02:04
Air Blank	0.000	02:05
Control Test	0.080	02:05
Air Blank	0.000	02:05
Subject Sample # 1	0.000	02:06
Breath Volume	1.140	02:06
Air Blank	0.000	02:06
Air Blank	0.000	02:08
Subject Sample # 2	0.000	02:09
Breath Volume	1.476	02:09
Air Blank	0.000	02:10
Control Test	0.081	02:10
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

---

Data Download Date/Time: 08/14/2019 09:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH, JOSIAH, R  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: BAIR, JAMES  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: LONDON

DATE/TIME OF TEST: 08/14/2019 20:59:18

---

Diagnostics Check	OK	20:53
Air Blank	0.000	20:53
Control Test	0.081	20:53
Air Blank	0.000	20:54
Subject Sample # 1	0.145	20:54
Breath Volume	2.046	20:54
Air Blank	0.000	20:55
Air Blank	0.000	20:57
Subject Sample # 2	0.145	20:57
Breath Volume	2.074	20:57
Air Blank	0.000	20:58
Control Test	0.079	20:58
Air Blank	0.000	20:59
Diagnostics Check	OK	20:59

---

Data Download Date/Time: 09/18/2019 10:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BONSALL, MATTHEW, A  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: COLON, Y  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: SOTO VALDEZ

DATE/TIME OF TEST: 08/17/2019 02:25:30

---

Diagnostics Check	OK	02:17
Air Blank	0.000	02:18
Control Test	0.080	02:18
Air Blank	0.000	02:19
Subject Sample # 1	0.102	02:20
Breath Volume	1.687	02:20
Air Blank	0.000	02:21
Air Blank	0.000	02:23
Subject Sample # 2	0.100	02:23
Breath Volume	1.566	02:23
Air Blank	0.000	02:24
Control Test	0.079	02:24
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

---

Data Download Date/Time: 09/18/2019 10:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BONSALL, MATTHEW, A  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: LAIRD, D  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: REYES HERNANDEZ

DATE/TIME OF TEST: 08/17/2019 02:42:45

---

Diagnostics Check		OK	02:31	
Air Blank		0.000	02:31	
Control Test		0.080	02:31	
Air Blank		0.000	02:32	
Subject Sample	# 1	0.218	02:35	Volume Not Met
Breath Volume		0.597	02:35	
Air Blank		0.000	02:36	
Air Blank		0.000	02:37	
Subject Sample	# 2	0.199	02:41	Volume Not Met
Breath Volume		0.000	02:41	
Air Blank		0.000	02:41	
Control Test		0.078	02:42	
Air Blank		0.000	02:42	
Diagnostics Check		OK	02:42	

---

Data Download Date/Time: 09/18/2019 10:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOULD,LEONARD,  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: GOULD,LEONARD  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 17:31  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 08/18/2019 19:29:24

---

Diagnostics Check	OK	19:23
Air Blank	0.000	19:23
Control Test	0.080	19:24
Air Blank	0.000	19:24
Subject Sample # 1	0.000	19:25
Breath Volume	1.152	19:25
Air Blank	0.000	19:25
Air Blank	0.000	19:27
Subject Sample # 2	0.000	19:28
Breath Volume	1.382	19:28
Air Blank	0.000	19:28
Control Test	0.080	19:28
Air Blank	0.000	19:29
Diagnostics Check	OK	19:29

---

Data Download Date/Time: 09/18/2019 10:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHILKE,CARL,E  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: SCHILKE,CARL  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: ORSTROM

DATE/TIME OF TEST: 08/19/2019 02:37:41

---

Diagnostics Check	OK	02:30
Air Blank	0.000	02:31
Control Test	0.080	02:31
Air Blank	0.000	02:31
Subject Sample # 1	0.210	02:32
Breath Volume	2.343	02:32
Air Blank	0.000	02:33
Air Blank	0.000	02:35
Subject Sample # 2	0.215	02:35
Breath Volume	4.203	02:35
Air Blank	0.000	02:36
Control Test	0.080	02:37
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

---

Data Download Date/Time: 09/18/2019 10:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WITT,KELLY,  
BREATH TEST OPERATOR AGENCY: Fort Myers PD  
ARRESTING OFFICER: WITT,KELLY  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:34  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: WIMBERLY

DATE/TIME OF TEST: 08/20/2019 00:34:59

---

Diagnostics Check	OK	00:27
Air Blank	0.000	00:28
Control Test	0.081	00:28
Air Blank	0.000	00:28
Subject Sample # 1	0.000	00:29
Breath Volume	2.269	00:29
Air Blank	0.000	00:30
Air Blank	0.000	00:32
Subject Sample # 2	0.000	00:33
Breath Volume	1.734	00:33
Air Blank	0.000	00:34
Control Test	0.080	00:34
Air Blank	0.000	00:34
Diagnostics Check	OK	00:34

---

Data Download Date/Time: 09/18/2019 10:04



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOULD,LEONARD,  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: GOULD,LEONARD  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 17:03  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 17:55

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 08/21/2019 18:21:27

---

Diagnostics Check	OK	18:15	
Air Blank	0.000	18:15	
Control Test	0.081	18:15	
Air Blank	0.000	18:16	
Subject Sample	# 1 0.053	18:19	Volume Not Met
Breath Volume	0.527	18:19	
Air Blank	0.000	18:20	
Air Blank	0.000	18:21	

---

Data Download Date/Time: 09/18/2019 10:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LINDHOLM,JOHN,T  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: LINDHOLM,JOHN  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: VARELA-CASTILLO

DATE/TIME OF TEST: 08/22/2019 02:36:25

---

Diagnostics Check	OK	02:28
Air Blank	0.000	02:29
Control Test	0.081	02:29
Air Blank	0.000	02:30
Subject Sample # 1	0.023	02:30
Breath Volume	1.824	02:30
Air Blank	0.000	02:31
Air Blank	0.000	02:33
Subject Sample # 2	0.024	02:34
Breath Volume	3.074	02:34
Air Blank	0.000	02:35
Control Test	0.082	02:35
Air Blank	0.000	02:36
Diagnostics Check	OK	02:36

---

Data Download Date/Time: 09/18/2019 10:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMYRNIOS,MANUEL,J  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: SMYRNIOS,MANUEL  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 10:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:25

SUBJECT'S LAST NAME: ZORILO

DATE/TIME OF TEST: 08/22/2019 11:52:08

---

Diagnostics Check	OK	11:45
Air Blank	0.000	11:46
Control Test	0.081	11:46
Air Blank	0.000	11:47
Subject Sample # 1	0.000	11:47
Breath Volume	2.667	11:47
Air Blank	0.000	11:48
Air Blank	0.000	11:50
Subject Sample # 2	0.000	11:50
Breath Volume	3.281	11:50
Air Blank	0.000	11:51
Control Test	0.080	11:51
Air Blank	0.000	11:52
Diagnostics Check	OK	11:52

---

Data Download Date/Time: 09/18/2019 10:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PERRY, RONALD, G  
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)  
ARRESTING OFFICER: GRAY, SETH  
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)  
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:47

SUBJECT'S LAST NAME: FARMER

DATE/TIME OF TEST: 08/22/2019 21:14:16

---

Diagnostics Check	OK	21:08	
Air Blank	0.000	21:08	
Control Test	0.080	21:09	
Air Blank	0.000	21:09	
Subject Sample # 1	0.112	21:10	
Breath Volume	3.457	21:10	
Air Blank	0.000	21:10	
Air Blank	0.000	21:12	
Subject Sample # 2	0.104	21:13	
Breath Volume	1.515	21:13	
Air Blank		21:13	RFI Detect
Air Blank	0.000	21:14	

---

Data Download Date/Time: 09/18/2019 10:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PERRY, RONALD, G  
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)  
ARRESTING OFFICER: GRAY, SETH  
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)  
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:47

SUBJECT'S LAST NAME: FARMER

DATE/TIME OF TEST: 08/22/2019 21:24:20

---

Diagnostics Check	OK	21:17
Air Blank	0.000	21:18
Control Test	0.079	21:18
Air Blank	0.000	21:19
Subject Sample # 1	0.097	21:19
Breath Volume	1.347	21:19
Air Blank	0.000	21:20
Air Blank	0.000	21:22
Subject Sample # 2	0.099	21:22
Breath Volume	1.648	21:22
Air Blank	0.000	21:23
Control Test	0.080	21:23
Air Blank	0.000	21:24
Diagnostics Check	OK	21:24

---

Data Download Date/Time: 09/18/2019 10:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TISDALE,KATIE,E  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: BREDECK,JORDAN  
ARRESTING OFFICER AGENCY: Lee Port Authority  
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:27  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: MIGLIORESE COLLARD

DATE/TIME OF TEST: 08/24/2019 02:16:30

---

Diagnostics Check		OK	02:10
Air Blank		0.000	02:10
Control Test		0.080	02:10
Air Blank		0.000	02:11
Subject Sample	# 1	0.314	02:11
Breath Volume		1.304	02:11
Air Blank		0.000	02:12
Air Blank		0.000	02:14
Subject Sample	# 2	0.303	02:14
Breath Volume		1.433	02:14
Air Blank		0.000	02:15
Control Test		0.079	02:15
Air Blank		0.000	02:16
Diagnostics Check		OK	02:16

---

Data Download Date/Time: 09/18/2019 10:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AVENDANO,JUAN,M  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: SIN,PHILLIP  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:25  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: CINTRON

DATE/TIME OF TEST: 08/24/2019 23:40:29

---

Diagnostics Check	OK	23:32
Air Blank	0.000	23:32
Control Test	0.081	23:33
Air Blank	0.000	23:33
Subject Sample # 1	0.161	23:35
Breath Volume	1.367	23:35
Air Blank	0.000	23:36
Air Blank	0.000	23:38
Subject Sample # 2	0.158	23:38
Breath Volume	1.335	23:38
Air Blank	0.000	23:39
Control Test	0.079	23:39
Air Blank	0.000	23:40
Diagnostics Check	OK	23:40

---

Data Download Date/Time: 09/18/2019 10:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOULD,LEONARD,  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: GOULD,LEONARD  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: SAGE

DATE/TIME OF TEST: 08/27/2019 23:07:09

---

Diagnostics Check	OK	23:01
Air Blank	0.000	23:01
Control Test	0.080	23:01
Air Blank	0.000	23:02
Subject Sample # 1	0.136	23:02
Breath Volume	2.582	23:02
Air Blank	0.000	23:03
Air Blank	0.000	23:05
Subject Sample # 2	0.134	23:05
Breath Volume	2.167	23:05
Air Blank	0.000	23:06
Control Test	0.081	23:06
Air Blank	0.000	23:07
Diagnostics Check	OK	23:07

---

Data Download Date/Time: 09/18/2019 10:04



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AVENDANO,JUAN,M  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: EUBANKS,AARON  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 15:49  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 17:11

SUBJECT'S LAST NAME: NARAYAN

DATE/TIME OF TEST: 08/29/2019 17:49:32

---

Diagnostics Check	OK	17:41
Air Blank	0.000	17:41
Control Test	0.081	17:42
Air Blank	0.000	17:42
Subject Sample # 1	0.237	17:43
Breath Volume	1.132	17:43
Air Blank	0.000	17:44
Air Blank	0.000	17:46
Subject Sample # 2	0.225	17:47
Breath Volume	1.281	17:47
Air Blank	0.000	17:48
Control Test	0.080	17:48
Air Blank	0.000	17:49
Diagnostics Check	OK	17:49

---

Data Download Date/Time: 09/18/2019 10:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHILKE,CARL,E  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: SCHILKE,CARL  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: BENITEZ FLORES

DATE/TIME OF TEST: 08/31/2019 00:23:22

---

Diagnostics Check		OK	00:15
Air Blank		0.000	00:15
Control Test		0.080	00:16
Air Blank		0.000	00:16
Subject Sample	# 1	0.121	00:17
Breath Volume		1.828	00:17
Air Blank		0.000	00:18
Air Blank		0.000	00:20
Subject Sample	# 2	0.125	00:21
Breath Volume		1.605	00:21
Air Blank		0.000	00:22
Control Test		0.081	00:22
Air Blank		0.000	00:23
Diagnostics Check		OK	00:23

---

Data Download Date/Time: 09/18/2019 10:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BONSALL, MATTHEW, A  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: BONSALL, M  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: BICKLEY

DATE/TIME OF TEST: 08/31/2019 03:41:12

---

Diagnostics Check	OK	03:34
Air Blank	0.000	03:35
Control Test	0.080	03:35
Air Blank	0.000	03:36
Subject Sample # 1	0.193	03:36
Breath Volume	1.957	03:36
Air Blank	0.000	03:37
Air Blank	0.000	03:38
Subject Sample # 2	0.195	03:39
Breath Volume	1.847	03:39
Air Blank	0.000	03:40
Control Test	0.081	03:40
Air Blank	0.000	03:41
Diagnostics Check	OK	03:41

---

Data Download Date/Time: 09/18/2019 10:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILSON,PERRY,M  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: WILSON,PERRY  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 07/19/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A1  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:41  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: MARTINEZ LOPEZ

DATE/TIME OF TEST: 08/04/2019 03:27:47

---

Diagnostics Check	OK	03:21
Air Blank	0.000	03:21
Control Test	0.080	03:22
Air Blank	0.000	03:22
Subject Sample # 1	0.111	03:23
Breath Volume	1.566	03:23
Air Blank	0.000	03:23
Air Blank	0.000	03:25
Subject Sample # 2	0.106	03:26
Breath Volume	1.878	03:26
Air Blank	0.000	03:26
Control Test	0.080	03:27
Air Blank	0.000	03:27
Diagnostics Check	OK	03:27

---

Data Download Date/Time: 08/14/2019 09:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARKLEY,JEFFREY,G  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: BARKLEY,JEFFREY  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 07/19/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A1  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 08/04/2019 03:37:03

---

Diagnostics Check	OK	03:30
Air Blank	0.000	03:30
Control Test	0.080	03:31
Air Blank	0.000	03:31
Subject Sample # 1	0.149	03:32
Breath Volume	1.500	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:34
Subject Sample # 2	0.151	03:35
Breath Volume	1.734	03:35
Air Blank	0.000	03:36
Control Test	0.079	03:36
Air Blank	0.000	03:36
Diagnostics Check	OK	03:37

---

Data Download Date/Time: 08/14/2019 09:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SATTERLEE, ANDREW, S  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: DOWDLE, OFFICER  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 07/19/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A1  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 12:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 12:44

SUBJECT'S LAST NAME: GOTKIEWICZ

DATE/TIME OF TEST: 08/05/2019 13:19:43

---

Diagnostics Check	OK	13:13
Air Blank	0.000	13:13
Control Test	0.080	13:13
Air Blank	0.000	13:14
Subject Sample # 1	0.000	13:15
Breath Volume	3.957	13:15
Air Blank	0.000	13:15
Air Blank	0.000	13:17
Subject Sample # 2	0.000	13:18
Breath Volume	2.457	13:18
Air Blank	0.000	13:18
Control Test	0.080	13:19
Air Blank	0.000	13:19
Diagnostics Check	OK	13:19

---

Data Download Date/Time: 08/14/2019 09:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MADURO,VICTOR,A  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: MADURO,VICTOR  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 07/19/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A1  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 15:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:14

SUBJECT'S LAST NAME: DEL RIO GARCI

DATE/TIME OF TEST: 08/06/2019 16:55:51

---

Diagnostics Check	OK	16:44	
Air Blank	0.000	16:44	
Control Test	0.080	16:45	
Air Blank	0.000	16:45	
Subject Sample	# 1	16:48	No Sample Provided
Breath Volume	0.000	16:48	
Air Blank	0.000	16:49	
Air Blank	0.000	16:51	
Subject Sample	# 2	16:54	No Sample Provided
Breath Volume	0.000	16:54	
Air Blank	0.000	16:54	
Control Test	0.079	16:55	
Air Blank	0.000	16:55	
Diagnostics Check	OK	16:55	

---

Data Download Date/Time: 08/14/2019 09:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARKLEY,JEFFREY,G  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: BARKLEY,JEFFREY  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 07/19/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A1  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: OTHER  
ARREST TIME: 14:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 08/07/2019 17:17:00

---

Diagnostics Check	OK	17:11
Air Blank	0.000	17:11
Control Test	0.080	17:12
Air Blank	0.000	17:12
Subject Sample # 1	0.000	17:12
Breath Volume	3.347	17:12
Air Blank	0.000	17:13
Air Blank	0.000	17:15
Subject Sample # 2	0.000	17:15
Breath Volume	3.585	17:15
Air Blank	0.000	17:16
Control Test	0.080	17:16
Air Blank	0.000	17:16
Diagnostics Check	OK	17:17

---

Data Download Date/Time: 08/14/2019 09:33



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARKLEY,JEFFREY,G  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: BARKLEY,JEFFREY  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 07/19/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A1  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: SCHUNDLEMIRE

DATE/TIME OF TEST: 08/10/2019 04:03:58

---

Diagnostics Check		OK	03:57
Air Blank		0.000	03:57
Control Test		0.080	03:58
Air Blank		0.000	03:58
Subject Sample	# 1	0.240	03:59
Breath Volume		2.003	03:59
Air Blank		0.000	04:00
Air Blank		0.000	04:01
Subject Sample	# 2	0.226	04:02
Breath Volume		1.386	04:02
Air Blank		0.000	04:03
Control Test		0.080	04:03
Air Blank		0.000	04:03
Diagnostics Check		OK	04:03

---

Data Download Date/Time: 08/14/2019 09:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILSON,PERRY,M  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: GRAY,DANNY  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 07/19/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A1  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:02  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: OQUENDO

DATE/TIME OF TEST: 08/11/2019 01:52:24

---

Diagnostics Check	OK	01:45
Air Blank	0.000	01:46
Control Test	0.079	01:46
Air Blank	0.000	01:47
Subject Sample # 1	0.155	01:47
Breath Volume	1.785	01:47
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample # 2	0.156	01:50
Breath Volume	3.246	01:50
Air Blank	0.000	01:51
Control Test	0.079	01:51
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

---

Data Download Date/Time: 08/14/2019 09:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILSON,PERRY,M  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: LOETHEN,KEITH  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 07/19/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A1  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 08/11/2019 02:03:22

---

Diagnostics Check	OK	01:56
Air Blank	0.000	01:57
Control Test	0.079	01:57
Air Blank	0.000	01:58
Subject Sample # 1	0.189	01:58
Breath Volume	2.144	01:58
Air Blank	0.000	01:59
Air Blank	0.000	02:01
Subject Sample # 2	0.184	02:01
Breath Volume	2.578	02:01
Air Blank	0.000	02:02
Control Test	0.079	02:02
Air Blank	0.000	02:03
Diagnostics Check	OK	02:03

---

Data Download Date/Time: 08/14/2019 09:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYERS,JUSTIN,L  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: MYERS,JUSTIN  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 07/19/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A1  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: HUERTAS

DATE/TIME OF TEST: 08/11/2019 03:48:21

---

Diagnostics Check	OK	03:40
Air Blank	0.000	03:41
Control Test	0.080	03:41
Air Blank	0.000	03:42
Subject Sample # 1	0.089	03:42
Breath Volume	1.367	03:42
Air Blank	0.000	03:43
Air Blank	0.000	03:45
Subject Sample # 2	0.084	03:46
Breath Volume	2.472	03:46
Air Blank	0.000	03:47
Control Test	0.080	03:47
Air Blank	0.000	03:48
Diagnostics Check	OK	03:48

---

Data Download Date/Time: 08/14/2019 09:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILSON,PERRY,M  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: WILSON,PERRY  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 07/19/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A1  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:24  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: LANDRY

DATE/TIME OF TEST: 08/11/2019 04:08:48

---

Diagnostics Check	OK	03:58
Air Blank	0.000	03:59
Control Test	0.079	03:59
Air Blank	0.000	04:00
Subject Sample # 1	0.202	04:02
Breath Volume	1.859	04:02
Air Blank	0.000	04:03
Air Blank	0.000	04:05
Subject Sample # 2	0.214	04:07
Breath Volume	1.660	04:07
Air Blank	0.000	04:07
Control Test	0.077	04:08
Air Blank	0.000	04:08
Diagnostics Check	OK	04:08

---

Data Download Date/Time: 08/14/2019 09:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTINI,RYAN,H  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: MARTINI,RYAN  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 07/19/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A1  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: KALKA

DATE/TIME OF TEST: 08/11/2019 22:48:51

---

Diagnostics Check	OK	22:42
Air Blank	0.000	22:42
Control Test	0.080	22:43
Air Blank	0.000	22:43
Subject Sample # 1	0.138	22:44
Breath Volume	1.574	22:44
Air Blank	0.000	22:44
Air Blank	0.000	22:46
Subject Sample # 2	0.135	22:47
Breath Volume	1.332	22:47
Air Blank	0.000	22:47
Control Test	0.079	22:48
Air Blank	0.000	22:48
Diagnostics Check	OK	22:48

---

Data Download Date/Time: 08/14/2019 09:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FILLAK, MICHAEL, O  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: SONG, GUANG  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A1  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: LEHMAN

DATE/TIME OF TEST: 08/16/2019 01:15:17

---

Diagnostics Check		OK	01:07
Air Blank		0.000	01:08
Control Test		0.080	01:08
Air Blank		0.000	01:08
Subject Sample	# 1	0.156	01:10
Breath Volume		4.941	01:10
Air Blank		0.000	01:10
Air Blank		0.000	01:12
Subject Sample	# 2	0.157	01:13
Breath Volume		4.882	01:13
Air Blank		0.000	01:14
Control Test		0.079	01:14
Air Blank		0.000	01:15
Diagnostics Check		OK	01:15

---

Data Download Date/Time: 09/18/2019 09:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ZARRILLO,N,  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: RUSH,D  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A1  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:57  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: HESSLER

DATE/TIME OF TEST: 08/17/2019 02:47:00

---

Diagnostics Check	OK	02:40
Air Blank	0.000	02:41
Control Test	0.080	02:41
Air Blank	0.000	02:42
Subject Sample # 1	0.213	02:42
Breath Volume	3.648	02:42
Air Blank	0.000	02:43
Air Blank	0.000	02:44
Subject Sample # 2	0.213	02:45
Breath Volume	3.558	02:45
Air Blank	0.000	02:46
Control Test	0.078	02:46
Air Blank	0.000	02:46
Diagnostics Check	OK	02:47

---

Data Download Date/Time: 09/18/2019 09:34



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYERS,JUSTIN,L  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: MYERS,JUSTIN  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A1  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:37  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: LOSO

DATE/TIME OF TEST: 08/17/2019 03:27:08

---

Diagnostics Check	OK	03:20
Air Blank	0.000	03:21
Control Test	0.080	03:21
Air Blank	0.000	03:21
Subject Sample # 1	0.100	03:22
Breath Volume	1.718	03:22
Air Blank	0.000	03:23
Air Blank	0.000	03:24
Subject Sample # 2	0.100	03:25
Breath Volume	2.371	03:25
Air Blank	0.000	03:26
Control Test	0.079	03:26
Air Blank	0.000	03:27
Diagnostics Check	OK	03:27

---

Data Download Date/Time: 09/18/2019 09:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYERS,JUSTIN,L  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: MYERS,JUSTIN  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A1  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: MARCELAK

DATE/TIME OF TEST: 08/18/2019 03:50:07

---

Diagnostics Check	OK	03:43
Air Blank	0.000	03:43
Control Test	0.080	03:44
Air Blank	0.000	03:44
Subject Sample # 1	0.119	03:45
Breath Volume	3.386	03:45
Air Blank	0.000	03:45
Air Blank	0.000	03:47
Subject Sample # 2	0.111	03:48
Breath Volume	2.257	03:48
Air Blank	0.000	03:49
Control Test	0.079	03:49
Air Blank	0.000	03:50
Diagnostics Check	OK	03:50

---

Data Download Date/Time: 09/18/2019 09:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DUFFY,J,  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: COLWELL,C  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A1  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:31  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: BALLARD

DATE/TIME OF TEST: 08/19/2019 01:24:32

---

Diagnostics Check	OK	01:18
Air Blank	0.000	01:18
Control Test	0.080	01:18
Air Blank	0.000	01:19
Subject Sample # 1	0.188	01:19
Breath Volume	1.914	01:19
Air Blank	0.000	01:20
Air Blank	0.000	01:22
Subject Sample # 2	0.193	01:22
Breath Volume	2.894	01:22
Air Blank	0.000	01:23
Control Test	0.080	01:23
Air Blank	0.000	01:24
Diagnostics Check	OK	01:24

---

Data Download Date/Time: 09/18/2019 09:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROSS,CHEYANNE,K  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: LOETHEN,KEITH  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A1  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: FOY

DATE/TIME OF TEST: 08/25/2019 03:00:58

---

Diagnostics Check	OK	02:54
Air Blank	0.000	02:54
Control Test	0.079	02:54
Air Blank	0.000	02:55
Subject Sample # 1	0.228	02:56
Breath Volume	3.863	02:56
Air Blank	0.000	02:56
Air Blank	0.000	02:58
Subject Sample # 2	0.236	02:59
Breath Volume	4.113	02:59
Air Blank	0.000	03:00
Control Test	0.079	03:00
Air Blank	0.000	03:00
Diagnostics Check	OK	03:00

---

Data Download Date/Time: 09/18/2019 09:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYERS,JUSTIN,L  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: FRAZIN,SEAN  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A1  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: SIEBERT

DATE/TIME OF TEST: 08/27/2019 02:29:09

---

Diagnostics Check	OK	02:22
Air Blank	0.000	02:23
Control Test	0.080	02:23
Air Blank	0.000	02:23
Subject Sample # 1	0.057	02:24
Breath Volume	4.312	02:24
Air Blank	0.000	02:25
Air Blank	0.000	02:26
Subject Sample # 2	0.056	02:27
Breath Volume	4.628	02:27
Air Blank	0.000	02:28
Control Test	0.080	02:28
Air Blank	0.000	02:29
Diagnostics Check	OK	02:29

---

Data Download Date/Time: 09/18/2019 09:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DUFFY,J,  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: DURHAM,K  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A1  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:13  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:36

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 08/27/2019 21:02:31

---

Diagnostics Check	OK	20:56
Air Blank	0.000	20:56
Control Test	0.080	20:57
Air Blank	0.000	20:57
Subject Sample # 1	0.000	20:58
Breath Volume	1.585	20:58
Air Blank	0.000	20:58
Air Blank	0.000	21:00
Subject Sample # 2	0.000	21:01
Breath Volume	2.511	21:01
Air Blank	0.000	21:01
Control Test	0.080	21:01
Air Blank	0.000	21:02
Diagnostics Check	OK	21:02

---

Data Download Date/Time: 09/18/2019 09:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYERS,JUSTIN,L  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: MYERS,JUSTIN  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A1  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:53  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:23

SUBJECT'S LAST NAME: PITTMAN

DATE/TIME OF TEST: 08/27/2019 21:50:30

---

Diagnostics Check	OK	21:43
Air Blank	0.000	21:44
Control Test	0.080	21:44
Air Blank	0.000	21:45
Subject Sample # 1	0.350	21:45
Breath Volume	1.984	21:45
Air Blank	0.000	21:46
Air Blank	0.000	21:48
Subject Sample # 2	0.352	21:48
Breath Volume	1.917	21:48
Air Blank	0.000	21:49
Control Test	0.079	21:49
Air Blank	0.000	21:50
Diagnostics Check	OK	21:50

---

Data Download Date/Time: 09/18/2019 09:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYERS,JUSTIN,L  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: SHIPPY,DYLAN  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A1  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 04:26  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 08/28/2019 05:02:46

---

Diagnostics Check	OK	04:55
Air Blank	0.000	04:55
Control Test	0.080	04:56
Air Blank	0.000	04:56
Subject Sample # 1	0.177	04:57
Breath Volume	1.886	04:57
Air Blank	0.000	04:58
Air Blank	0.000	05:00
Subject Sample # 2	0.179	05:01
Breath Volume	1.648	05:01
Air Blank	0.000	05:01
Control Test	0.080	05:02
Air Blank	0.000	05:02
Diagnostics Check	OK	05:02

---

Data Download Date/Time: 09/18/2019 09:34



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTINI,RYAN,H  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: VENSKI,JASON  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A1  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:53

SUBJECT'S LAST NAME: QUINTILIANI

DATE/TIME OF TEST: 08/29/2019 21:21:25

---

Diagnostics Check	OK	21:13
Air Blank	0.000	21:14
Control Test	0.080	21:14
Air Blank	0.000	21:15
Subject Sample # 1	0.062	21:16
Breath Volume	2.527	21:16
Air Blank	0.000	21:17
Air Blank	0.000	21:19
Subject Sample # 2	0.061	21:19
Breath Volume	2.988	21:19
Air Blank	0.000	21:20
Control Test	0.080	21:20
Air Blank	0.000	21:21
Diagnostics Check	OK	21:21

---

Data Download Date/Time: 09/18/2019 09:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTINI,RYAN,H  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: MEADOWS,KELSEY  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A1  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: ARMSTRONG

DATE/TIME OF TEST: 08/29/2019 21:30:29

---

Diagnostics Check	OK	21:24
Air Blank	0.000	21:24
Control Test	0.080	21:24
Air Blank	0.000	21:25
Subject Sample # 1	0.226	21:25
Breath Volume	1.988	21:25
Air Blank	0.000	21:26
Air Blank	0.000	21:28
Subject Sample # 2	0.231	21:28
Breath Volume	1.933	21:28
Air Blank	0.000	21:29
Control Test	0.079	21:29
Air Blank	0.000	21:30
Diagnostics Check	OK	21:30

---

Data Download Date/Time: 09/18/2019 09:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AMATO,MARK,P  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: MONTGOMERY,JOHN  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A1  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 07:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:34

SUBJECT'S LAST NAME: BOYD

DATE/TIME OF TEST: 08/31/2019 08:01:05

---

Diagnostics Check	OK	07:54
Air Blank	0.000	07:55
Control Test	0.080	07:55
Air Blank	0.000	07:56
Subject Sample # 1	0.142	07:56
Breath Volume	1.187	07:56
Air Blank	0.000	07:57
Air Blank	0.000	07:59
Subject Sample # 2	0.134	07:59
Breath Volume	1.695	07:59
Air Blank	0.000	08:00
Control Test	0.079	08:00
Air Blank	0.000	08:01
Diagnostics Check	OK	08:01

---

Data Download Date/Time: 09/18/2019 09:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCCLELLAND, WILLIAM, R  
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO  
ARRESTING OFFICER: MILLER, K  
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO  
LAST AGENCY INSPECTION: 08/01/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:48

SUBJECT'S LAST NAME: RUIZ

DATE/TIME OF TEST: 08/13/2019 21:40:27

---

Diagnostics Check	OK	21:34
Air Blank	0.000	21:34
Control Test	0.078	21:34
Air Blank	0.000	21:35
Subject Sample # 1	0.191	21:35
Breath Volume	3.093	21:35
Air Blank	0.000	21:36
Air Blank	0.000	21:38
Subject Sample # 2	0.190	21:38
Breath Volume	2.925	21:38
Air Blank	0.000	21:39
Control Test	0.078	21:39
Air Blank	0.000	21:40
Diagnostics Check	OK	21:40

---

Data Download Date/Time: 09/04/2019 14:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WESTON,LARRY,L  
BREATH TEST OPERATOR AGENCY: Glades County SO  
ARRESTING OFFICER: MURRAY,STACY  
ARRESTING OFFICER AGENCY: Glades County SO  
LAST AGENCY INSPECTION: 07/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: ADMIN  
ARREST TIME: 23:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: COLON

DATE/TIME OF TEST: 08/03/2019 07:06:16

---

Diagnostics Check		OK	06:59
Air Blank		0.000	07:00
Control Test		0.077	07:00
Air Blank		0.000	07:00
Subject Sample	# 1	0.125	07:01
Breath Volume		1.300	07:01
Air Blank		0.000	07:02
Air Blank		0.000	07:03
Subject Sample	# 2	0.124	07:04
Breath Volume		1.710	07:04
Air Blank		0.000	07:05
Control Test		0.076	07:05
Air Blank		0.000	07:06
Diagnostics Check		OK	07:06

---

Data Download Date/Time: 09/12/2019 19:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WESTON,LARRY,L  
BREATH TEST OPERATOR AGENCY: Glades County SO  
ARRESTING OFFICER: MURRAY,STACY  
ARRESTING OFFICER AGENCY: Glades County SO  
LAST AGENCY INSPECTION: 07/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: ADMIN  
ARREST TIME: 23:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: COLON

DATE/TIME OF TEST: 08/03/2019 10:38:01

---

Diagnostics Check	OK	10:31
Air Blank	0.000	10:31
Control Test	0.077	10:32
Air Blank	0.000	10:32
Subject Sample # 1	0.058	10:32
Breath Volume	1.339	10:32
Air Blank	0.000	10:33
Air Blank	0.000	10:35
Subject Sample # 2	0.061	10:36
Breath Volume	1.390	10:36
Air Blank	0.000	10:37
Control Test	0.077	10:37
Air Blank	0.000	10:37
Diagnostics Check	OK	10:38

---

Data Download Date/Time: 09/12/2019 19:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MADRUGO,FELICIA,  
BREATH TEST OPERATOR AGENCY: Glades County SO  
ARRESTING OFFICER: CLENDENING,JAMES  
ARRESTING OFFICER AGENCY: Glades County SO  
LAST AGENCY INSPECTION: 07/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: WALTERS

DATE/TIME OF TEST: 08/14/2019 21:04:39

---

Diagnostics Check		OK	20:58
Air Blank		0.000	20:58
Control Test		0.076	20:59
Air Blank		0.000	20:59
Subject Sample	# 1	0.206	21:00
Breath Volume		1.164	21:00
Air Blank		0.000	21:00
Air Blank		0.000	21:02
Subject Sample	# 2	0.210	21:03
Breath Volume		1.183	21:03
Air Blank		0.000	21:03
Control Test		0.076	21:04
Air Blank		0.000	21:04
Diagnostics Check		OK	21:04

---

Data Download Date/Time: 09/12/2019 19:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MADRUGO,FELICIA,  
BREATH TEST OPERATOR AGENCY: Glades County SO  
ARRESTING OFFICER: CLENDENING,JAMES  
ARRESTING OFFICER AGENCY: Glades County SO  
LAST AGENCY INSPECTION: 07/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: WALTERS

DATE/TIME OF TEST: 08/15/2019 04:42:32

---

Diagnostics Check	OK	04:36
Air Blank	0.000	04:36
Control Test	0.077	04:37
Air Blank	0.000	04:37
Subject Sample # 1	0.064	04:38
Breath Volume	1.105	04:38
Air Blank	0.000	04:38
Air Blank	0.000	04:40
Subject Sample # 2	0.063	04:40
Breath Volume	1.121	04:40
Air Blank	0.000	04:41
Control Test	0.077	04:41
Air Blank	0.000	04:42
Diagnostics Check	OK	04:42

---

Data Download Date/Time: 09/12/2019 19:25



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LIVINGSTON,JESSICA,L  
BREATH TEST OPERATOR AGENCY: Glades County SO  
ARRESTING OFFICER: HERNANDEZ,DANIEL  
ARRESTING OFFICER AGENCY: Glades County SO  
LAST AGENCY INSPECTION: 07/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 08/18/2019 23:03:42

---

Diagnostics Check	OK	22:56
Air Blank	0.000	22:57
Control Test	0.078	22:57
Air Blank	0.000	22:58
Subject Sample # 1	0.201	22:59
Breath Volume	2.906	22:59
Air Blank	0.000	22:59
Air Blank	0.000	23:01
Subject Sample # 2	0.207	23:01
Breath Volume	2.457	23:01
Air Blank	0.000	23:02
Control Test	0.077	23:03
Air Blank	0.000	23:03
Diagnostics Check	OK	23:03

---

Data Download Date/Time: 09/12/2019 19:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EPPS,QUINTELLA,  
BREATH TEST OPERATOR AGENCY: Glades County SO  
ARRESTING OFFICER: HERNANDEZ,JOEL  
ARRESTING OFFICER AGENCY: Glades County SO  
LAST AGENCY INSPECTION: 07/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: GUERRERO GARCIA

DATE/TIME OF TEST: 08/24/2019 03:04:22

---

Diagnostics Check		OK	02:51	
Air Blank		0.000	02:52	
Control Test		0.077	02:52	
Air Blank		0.000	02:52	
Subject Sample	# 1	0.159	02:56	Volume Not Met
Breath Volume		0.140	02:56	
Air Blank		0.000	02:56	
Air Blank		0.000	02:58	
Subject Sample	# 2	0.161	02:59	
Breath Volume		1.664	02:59	
Air Blank		0.000	03:00	
Air Blank		0.000	03:02	
Subject Sample	# 3	0.171	03:02	
Breath Volume		1.269	03:02	
Air Blank		0.000	03:03	
Control Test		0.074	03:03	Control Outside Tolerance
Air Blank		0.000	03:04	

---

Data Download Date/Time: 09/12/2019 19:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EPPS,QUINTELLA,  
BREATH TEST OPERATOR AGENCY: Glades County SO  
ARRESTING OFFICER: CLENDEING,JAMES  
ARRESTING OFFICER AGENCY: Glades County SO  
LAST AGENCY INSPECTION: 07/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 18:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:48

SUBJECT'S LAST NAME: BURGOS

DATE/TIME OF TEST: 08/25/2019 19:35:06

---

Diagnostics Check	OK	19:26
Air Blank	0.000	19:26
Control Test	0.077	19:27
Air Blank	0.000	19:27
Subject Sample # 1	0.187	19:29
Breath Volume	1.652	19:29
Air Blank	0.000	19:30
Air Blank	0.000	19:31
Subject Sample # 2	0.180	19:33
Breath Volume	1.351	19:33
Air Blank	0.000	19:34
Control Test	0.075	19:34
Air Blank	0.000	19:35
Diagnostics Check	OK	19:35

---

Data Download Date/Time: 09/12/2019 19:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000949

INSTRUMENT REGISTERED TO: TALLAHASSEE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KELLER,DAVID,  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: KELLER,DAVID  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 07/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A1  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: LENNIE

DATE/TIME OF TEST: 08/09/2019 03:08:06

---

Diagnostics Check	OK	02:58
Air Blank	0.000	02:59
Control Test	0.078	02:59
Air Blank	0.000	03:00
Subject Sample # 1	0.206	03:00
Breath Volume	2.753	03:00
Air Blank	0.000	03:01
Air Blank	0.000	03:03
Subject Sample # 2	0.182	03:03
Breath Volume	1.226	03:03
Air Blank	0.000	03:04
Air Blank	0.000	03:06
Subject Sample # 3	0.208	03:06
Breath Volume	2.175	03:06
Air Blank	0.000	03:07
Control Test	0.080	03:07
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

---

Data Download Date/Time: 08/22/2019 16:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000954

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FINUFF,DANIEL,W  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: MCCOY,DUSTIN  
ARRESTING OFFICER AGENCY: Jefferson County SO  
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29017080A3  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: CARLSON

DATE/TIME OF TEST: 08/18/2019 23:02:34

---

Diagnostics Check		OK	22:56
Air Blank		0.000	22:56
Control Test		0.077	22:57
Air Blank		0.000	22:57
Subject Sample	# 1	0.308	22:58
Breath Volume		2.992	22:58
Air Blank		0.000	22:58
Air Blank		0.000	23:00
Subject Sample	# 2	0.301	23:01
Breath Volume		4.292	23:01
Air Blank		0.000	23:01
Control Test		0.076	23:01
Air Blank		0.000	23:02
Diagnostics Check		OK	23:02

---

Data Download Date/Time: 09/09/2019 13:38

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000954

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH, DAVID, H  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: ADAMCZYK, ANDREW  
ARRESTING OFFICER AGENCY: Jefferson County SO  
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29017080A3  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 08/27/2019 22:32:12

---

Diagnostics Check	OK	22:25
Air Blank	0.000	22:26
Control Test	0.077	22:26
Air Blank	0.000	22:26
Subject Sample # 1	0.054	22:27
Breath Volume	1.785	22:27
Air Blank	0.000	22:28
Air Blank	0.000	22:30
Subject Sample # 2	0.052	22:30
Breath Volume	3.066	22:30
Air Blank	0.000	22:31
Control Test	0.077	22:31
Air Blank	0.000	22:32
Diagnostics Check	OK	22:32

---

Data Download Date/Time: 09/09/2019 13:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: EPPS,TIMOTHY  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 07/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:23  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 08/02/2019 01:47:05

---

Diagnostics Check	OK	01:41
Air Blank	0.000	01:41
Control Test	0.082	01:41
Air Blank	0.000	01:42
Subject Sample # 1	0.138	01:42
Breath Volume	3.109	01:42
Air Blank	0.000	01:43
Air Blank	0.000	01:45
Subject Sample # 2	0.125	01:45
Breath Volume	1.246	01:45
Air Blank	0.000	01:46
Control Test	0.081	01:46
Air Blank	0.000	01:46
Diagnostics Check	OK	01:47

---

Data Download Date/Time: 08/07/2019 14:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARDEN,WILLIAM,L  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: VENTURA,CHRISTOPHER  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 07/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 18:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:48

SUBJECT'S LAST NAME: RUIZ-RUIZ

DATE/TIME OF TEST: 08/04/2019 20:25:43

---

Diagnostics Check	OK	20:17
Air Blank	0.000	20:17
Control Test	0.082	20:18
Air Blank	0.000	20:18
Subject Sample # 1	0.174	20:20
Breath Volume	1.421	20:20
Air Blank	0.000	20:21
Air Blank	0.000	20:23
Subject Sample # 2	0.181	20:24
Breath Volume	1.562	20:24
Air Blank	0.000	20:24
Control Test	0.080	20:25
Air Blank	0.000	20:25
Diagnostics Check	OK	20:25

---

Data Download Date/Time: 08/07/2019 14:43



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARDEN,WILLIAM,L  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: HOLLEY,DEANDRE  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 07/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: HERNANDEZ-LUCERO

DATE/TIME OF TEST: 08/05/2019 21:04:57

---

Diagnostics Check	OK	20:57
Air Blank	0.000	20:57
Control Test	0.081	20:58
Air Blank	0.000	20:58
Subject Sample # 1	0.172	20:59
Breath Volume	1.722	20:59
Air Blank	0.000	20:59
Air Blank	0.000	21:01
Subject Sample # 2	0.180	21:03
Breath Volume	2.617	21:03
Air Blank	0.000	21:04
Control Test	0.081	21:04
Air Blank	0.000	21:04
Diagnostics Check	OK	21:04

---

Data Download Date/Time: 08/07/2019 14:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TIAMSON,DAWN,P  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: JORGE,JUNIOR  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 07/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: LOZIN

DATE/TIME OF TEST: 08/11/2019 01:00:43

---

Diagnostics Check		OK	00:54
Air Blank		0.000	00:54
Control Test		0.081	00:55
Air Blank		0.000	00:55
Subject Sample	# 1	0.218	00:56
Breath Volume		1.128	00:56
Air Blank		0.000	00:56
Air Blank		0.000	00:58
Subject Sample	# 2	0.220	00:59
Breath Volume		1.250	00:59
Air Blank		0.000	00:59
Control Test		0.081	01:00
Air Blank		0.000	01:00
Diagnostics Check		OK	01:00

---

Data Download Date/Time: 08/15/2019 12:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARDEN,WILLIAM,L  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: VINER,DENNIS  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 07/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: OROZCO

DATE/TIME OF TEST: 08/11/2019 21:29:17

---

Diagnostics Check	OK	21:17
Air Blank	0.000	21:18
Control Test	0.082	21:18
Air Blank	0.000	21:19
Subject Sample # 1	0.291	21:21
Breath Volume	1.515	21:21
Air Blank	0.000	21:22
Air Blank	0.000	21:24
Subject Sample # 2	0.264	21:24
Breath Volume	1.367	21:24
Air Blank	0.000	21:25
Air Blank	0.000	21:27
Subject Sample # 3	0.256	21:27
Breath Volume	1.382	21:27
Air Blank	0.000	21:28
Control Test	0.080	21:28
Air Blank	0.000	21:29
Diagnostics Check	OK	21:29

---

Data Download Date/Time: 08/15/2019 12:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: HARRIS,BRIAN  
ARRESTING OFFICER AGENCY: Ocoee PD  
LAST AGENCY INSPECTION: 08/15/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 04:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: SAUCEDO

DATE/TIME OF TEST: 08/17/2019 05:11:29

---

Diagnostics Check	OK	05:05
Air Blank	0.000	05:05
Control Test	0.082	05:06
Air Blank	0.000	05:06
Subject Sample # 1	0.103	05:07
Breath Volume	1.546	05:07
Air Blank	0.000	05:07
Air Blank	0.000	05:09
Subject Sample # 2	0.104	05:09
Breath Volume	2.476	05:09
Air Blank	0.000	05:10
Control Test	0.081	05:10
Air Blank	0.000	05:11
Diagnostics Check	OK	05:11

---

Data Download Date/Time: 08/22/2019 13:46

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: MENDOZA, ANGEL  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 08/15/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: BUCKLEY

DATE/TIME OF TEST: 08/18/2019 02:39:39

---

Diagnostics Check	OK	02:36	
Air Blank	0.000	02:36	
Control Test	0.082	02:37	
Air Blank	0.000	02:37	
Subject Sample # 1	0.166	02:38	
Breath Volume	2.046	02:38	
Air Blank		02:38	RFI Detect
Air Blank	0.000	02:38	
Control Test	0.081	02:39	
Air Blank		02:39	RFI Detect
Air Blank	0.000	02:39	

---

Data Download Date/Time: 08/22/2019 13:46

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: MENDOZA,ANGEL  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 08/15/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: BUCKLEY

DATE/TIME OF TEST: 08/18/2019 02:44:19

---

Diagnostics Check	OK	02:41	
Air Blank	0.000	02:41	
Control Test	0.080	02:41	
Air Blank	0.000	02:42	
Subject Sample # 1	0.167	02:43	
Breath Volume	2.410	02:43	
Air Blank		02:43	RFI Detect
Air Blank		02:43	RFI Detect
Air Blank	0.000	02:44	

---

Data Download Date/Time: 08/22/2019 13:46

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: MOONDA, SHUAIB  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 08/15/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: SOTO

DATE/TIME OF TEST: 08/21/2019 02:21:45

---

Diagnostics Check	OK	02:15
Air Blank	0.000	02:16
Control Test	0.082	02:16
Air Blank	0.000	02:17
Subject Sample # 1	0.000	02:17
Breath Volume	1.726	02:17
Air Blank	0.000	02:18
Air Blank	0.000	02:20
Subject Sample # 2	0.000	02:20
Breath Volume	1.242	02:20
Air Blank	0.000	02:20
Control Test	0.081	02:21
Air Blank	0.000	02:21
Diagnostics Check	OK	02:21

---

Data Download Date/Time: 08/22/2019 13:46

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DENSON, TRAVIS, W  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: RIVERA, HERIBERTO  
ARRESTING OFFICER AGENCY: Winter Garden PD  
LAST AGENCY INSPECTION: 08/15/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: JAMIESON

DATE/TIME OF TEST: 08/23/2019 23:39:00

---

Diagnostics Check	OK	23:32
Air Blank	0.000	23:33
Control Test	0.082	23:33
Air Blank	0.000	23:34
Subject Sample # 1	0.000	23:34
Breath Volume	4.324	23:34
Air Blank	0.000	23:35
Air Blank	0.000	23:37
Subject Sample # 2	0.000	23:37
Breath Volume	4.437	23:37
Air Blank	0.000	23:38
Control Test	0.081	23:38
Air Blank	0.000	23:38
Diagnostics Check	OK	23:39

---

Data Download Date/Time: 08/29/2019 16:32



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: AUSTIN,ADAM  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 08/15/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 08/25/2019 03:16:53

---

Diagnostics Check	OK	03:08
Air Blank	0.000	03:09
Control Test	0.081	03:09
Air Blank	0.000	03:09
Subject Sample # 1	0.113	03:11
Breath Volume	1.261	03:11
Air Blank	0.000	03:11
Air Blank	0.000	03:13
Subject Sample # 2	0.105	03:15
Breath Volume	1.289	03:15
Air Blank	0.000	03:15
Control Test	0.080	03:16
Air Blank	0.000	03:16
Diagnostics Check	OK	03:16

---

Data Download Date/Time: 08/29/2019 16:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVIS,COURTNEY,L  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: AUSTIN,ADAM  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 08/15/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 08/31/2019 02:33:55

---

Diagnostics Check	OK	02:26
Air Blank	0.000	02:27
Control Test	0.082	02:27
Air Blank	0.000	02:28
Subject Sample # 1	0.087	02:29
Breath Volume	2.312	02:29
Air Blank	0.000	02:29
Air Blank	0.000	02:31
Subject Sample # 2	0.090	02:32
Breath Volume	2.558	02:32
Air Blank	0.000	02:33
Control Test	0.081	02:33
Air Blank	0.000	02:33
Diagnostics Check	OK	02:33

---

Data Download Date/Time: 09/04/2019 14:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DENSON, TRAVIS, W  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: MCFERREN, JASON  
ARRESTING OFFICER AGENCY: Maitland PD  
LAST AGENCY INSPECTION: 08/15/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: MCLAUGHLIN

DATE/TIME OF TEST: 08/31/2019 02:49:38

---

Diagnostics Check	OK	02:43
Air Blank	0.000	02:43
Control Test	0.081	02:43
Air Blank	0.000	02:44
Subject Sample # 1	0.119	02:45
Breath Volume	1.832	02:45
Air Blank	0.000	02:45
Air Blank	0.000	02:47
Subject Sample # 2	0.121	02:48
Breath Volume	1.496	02:48
Air Blank	0.000	02:48
Control Test	0.081	02:49
Air Blank	0.000	02:49
Diagnostics Check	OK	02:49

---

Data Download Date/Time: 09/04/2019 14:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVIS,COURTNEY,L  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: PARRISH,KAYLA  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 08/15/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 08/31/2019 04:20:29

---

Diagnostics Check		OK	04:12
Air Blank		0.000	04:13
Control Test		0.082	04:13
Air Blank		0.000	04:13
Subject Sample	# 1	0.034	04:15
Breath Volume		1.367	04:15
Air Blank		0.000	04:15
Air Blank		0.000	04:17
Subject Sample	# 2	0.034	04:19
Breath Volume		1.570	04:19
Air Blank		0.000	04:19
Control Test		0.081	04:19
Air Blank		0.000	04:20
Diagnostics Check		OK	04:20

---

Data Download Date/Time: 09/04/2019 14:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILKINS,CONNOR,R  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: IMPARATO,FRANK  
ARRESTING OFFICER AGENCY: Central FL Univ PD  
LAST AGENCY INSPECTION: 08/15/2019

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:03  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: GARS

DATE/TIME OF TEST: 08/31/2019 03:10:27

---

Diagnostics Check	OK	03:03
Air Blank	0.000	03:03
Control Test	0.079	03:03
Air Blank	0.000	03:04
Subject Sample # 1	0.258	03:05
Breath Volume	1.757	03:05
Air Blank	0.000	03:06
Air Blank	0.000	03:07
Subject Sample # 2	0.266	03:08
Breath Volume	2.277	03:08
Air Blank	0.000	03:09
Control Test	0.079	03:09
Air Blank	0.000	03:10
Diagnostics Check	OK	03:10

---

Data Download Date/Time: 09/04/2019 15:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TIAMSON,DAWN,P  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: WALCZAK,KEVIN  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 08/15/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: GUTIERREZ

DATE/TIME OF TEST: 08/31/2019 04:19:09

---

Diagnostics Check		OK	04:12
Air Blank		0.000	04:12
Control Test		0.079	04:12
Air Blank		0.000	04:13
Subject Sample	# 1	0.203	04:14
Breath Volume		2.367	04:14
Air Blank		0.000	04:15
Air Blank		0.000	04:17
Subject Sample	# 2	0.196	04:17
Breath Volume		1.195	04:17
Air Blank		0.000	04:18
Control Test		0.078	04:18
Air Blank		0.000	04:19
Diagnostics Check		OK	04:19

---

Data Download Date/Time: 09/04/2019 15:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARDEN,WILLIAM,L  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: PETERSON,CAMERON  
ARRESTING OFFICER AGENCY: Winter Garden PD  
LAST AGENCY INSPECTION: 08/15/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: AHEDO GUTIERREZ

DATE/TIME OF TEST: 08/31/2019 04:39:00

---

Diagnostics Check	OK	04:32
Air Blank	0.000	04:33
Control Test	0.079	04:33
Air Blank	0.000	04:34
Subject Sample # 1	0.146	04:34
Breath Volume	1.542	04:34
Air Blank	0.000	04:35
Air Blank	0.000	04:37
Subject Sample # 2	0.151	04:37
Breath Volume	1.503	04:37
Air Blank	0.000	04:38
Control Test	0.080	04:38
Air Blank	0.000	04:38
Diagnostics Check	OK	04:39

---

Data Download Date/Time: 09/04/2019 15:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMPSON,RONDRICKA,N  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: THOMPSON,RONDRICKA  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 07/15/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29017080A3  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:36  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: BRANNON

DATE/TIME OF TEST: 08/06/2019 01:40:30

---

Diagnostics Check	OK	01:29
Air Blank	0.000	01:29
Control Test	0.078	01:29
Air Blank	0.000	01:30
Subject Sample # 1	0.106	01:33
Breath Volume	1.101	01:33
Air Blank	0.000	01:34
Air Blank	0.000	01:35
Subject Sample # 2	0.102	01:38
Breath Volume	1.175	01:38
Air Blank	0.000	01:39
Control Test	0.077	01:39
Air Blank	0.000	01:40
Diagnostics Check	OK	01:40

---

Data Download Date/Time: 08/21/2019 10:15



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FINUFF,DANIEL,W  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: FINUFF,DANIEL  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 07/15/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29017080A3  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:34  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: POLLOCK

DATE/TIME OF TEST: 08/15/2019 02:21:00

---

Diagnostics Check	OK	02:14
Air Blank	0.000	02:14
Control Test	0.078	02:15
Air Blank	0.000	02:15
Subject Sample # 1	0.139	02:16
Breath Volume	2.644	02:16
Air Blank	0.000	02:16
Air Blank	0.000	02:18
Subject Sample # 2	0.128	02:19
Breath Volume	1.125	02:19
Air Blank	0.000	02:20
Control Test	0.079	02:20
Air Blank	0.000	02:20
Diagnostics Check	OK	02:21

---

Data Download Date/Time: 08/21/2019 10:15

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLARK,ASHLEY,J  
BREATH TEST OPERATOR AGENCY: Quincy PD  
ARRESTING OFFICER: CLARK,ASHLEY  
ARRESTING OFFICER AGENCY: Quincy PD  
LAST AGENCY INSPECTION: 07/15/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29017080A3  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: MILLS

DATE/TIME OF TEST: 08/17/2019 02:14:45

---

Diagnostics Check	OK	02:06
Air Blank	0.000	02:07
Control Test	0.077	02:07
Air Blank	0.000	02:08
Subject Sample # 1	0.127	02:09
Breath Volume	1.261	02:09
Air Blank	0.000	02:09
Air Blank	0.000	02:11
Subject Sample # 2	0.113	02:13
Breath Volume	1.347	02:13
Air Blank	0.000	02:13
Control Test	0.077	02:14
Air Blank	0.000	02:14
Diagnostics Check	OK	02:14

---

Data Download Date/Time: 08/21/2019 10:15

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GAINES,CRAIG,A  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: JETER,TRENTON  
ARRESTING OFFICER AGENCY: Chattahoochee PD  
LAST AGENCY INSPECTION: 08/21/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29017080A3  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 17:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:33

SUBJECT'S LAST NAME: COLLIER

DATE/TIME OF TEST: 08/21/2019 20:06:57

---

Diagnostics Check	OK	20:00
Air Blank	0.000	20:00
Control Test	0.078	20:00
Air Blank	0.000	20:01
Subject Sample # 1	0.170	20:02
Breath Volume	1.609	20:02
Air Blank	0.000	20:02
Air Blank	0.000	20:04
Subject Sample # 2	0.175	20:05
Breath Volume	1.914	20:05
Air Blank	0.000	20:06
Control Test	0.077	20:06
Air Blank	0.000	20:06
Diagnostics Check	OK	20:06

---

Data Download Date/Time: 09/11/2019 12:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FINUFF,DANIEL,W  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: FINUFF,DANIEL  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 08/21/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29017080A3  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 08/25/2019 00:54:22

---

Diagnostics Check	OK	00:51	
Air Blank	0.000	00:51	
Control Test	0.078	00:52	
Air Blank	0.000	00:52	
Subject Sample	# 1	00:53	Subject Test Refused
Breath Volume	0.000	00:53	
Air Blank	0.000	00:53	
Control Test	0.078	00:53	
Air Blank	0.000	00:54	
Diagnostics Check	OK	00:54	

---

Data Download Date/Time: 09/11/2019 12:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PYE,CHRISTOPHER,T  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: PYE,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 07/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:46  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 08/02/2019 00:23:40

---

Diagnostics Check	OK	00:17
Air Blank	0.000	00:17
Control Test	0.079	00:18
Air Blank	0.000	00:18
Subject Sample # 1	0.116	00:19
Breath Volume	1.648	00:19
Air Blank	0.000	00:19
Air Blank	0.000	00:21
Subject Sample # 2	0.119	00:22
Breath Volume	1.589	00:22
Air Blank	0.000	00:22
Control Test	0.079	00:23
Air Blank	0.000	00:23
Diagnostics Check	OK	00:23

---

Data Download Date/Time: 08/02/2019 19:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCCONNELL,CHARLES,E  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: MCCONNELL,CHARLES  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 08/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 14:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:15

SUBJECT'S LAST NAME: CARRION-DIAZ

DATE/TIME OF TEST: 08/03/2019 15:42:49

---

Diagnostics Check	OK	15:35
Air Blank	0.000	15:36
Control Test	0.079	15:36
Air Blank	0.000	15:36
Subject Sample # 1	0.000	15:38
Breath Volume	2.312	15:38
Air Blank	0.000	15:38
Air Blank	0.000	15:40
Subject Sample # 2	0.000	15:41
Breath Volume	2.082	15:41
Air Blank	0.000	15:41
Control Test	0.078	15:42
Air Blank	0.000	15:42
Diagnostics Check	OK	15:42

---

Data Download Date/Time: 09/11/2019 00:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PYE,CHRISTOPHER,T  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: PYE,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 08/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 20:42  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: RAMOS COLON

DATE/TIME OF TEST: 08/05/2019 21:25:33

---

Diagnostics Check	OK	21:18
Air Blank	0.000	21:19
Control Test	0.080	21:19
Air Blank	0.000	21:20
Subject Sample # 1	0.000	21:21
Breath Volume	2.222	21:21
Air Blank	0.000	21:21
Air Blank	0.000	21:23
Subject Sample # 2	0.000	21:24
Breath Volume	2.019	21:24
Air Blank	0.000	21:24
Control Test	0.078	21:24
Air Blank	0.000	21:25
Diagnostics Check	OK	21:25

---

Data Download Date/Time: 09/11/2019 00:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PYE,CHRISTOPHER,T  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: PYE,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 08/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: WORKMAN

DATE/TIME OF TEST: 08/06/2019 03:17:17

---

Diagnostics Check	OK	03:11
Air Blank	0.000	03:11
Control Test	0.079	03:11
Air Blank	0.000	03:12
Subject Sample # 1	0.031	03:12
Breath Volume	1.191	03:12
Air Blank	0.000	03:13
Air Blank	0.000	03:15
Subject Sample # 2	0.033	03:15
Breath Volume	2.066	03:15
Air Blank	0.000	03:16
Control Test	0.078	03:16
Air Blank	0.000	03:17
Diagnostics Check	OK	03:17

---

Data Download Date/Time: 09/11/2019 00:09



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LYONS,DANIEL,E  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: CRAWFORD,DAVID  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 08/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 20:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:27

SUBJECT'S LAST NAME: HAYNES

DATE/TIME OF TEST: 08/08/2019 21:16:52

---

Diagnostics Check	OK	21:13	
Air Blank	0.000	21:13	
Control Test	0.079	21:14	
Air Blank	0.000	21:14	
Subject Sample	# 1	21:15	Subject Test Refused
Breath Volume	0.000	21:15	
Air Blank	0.000	21:15	
Control Test	0.079	21:16	
Air Blank	0.000	21:16	
Diagnostics Check	OK	21:16	

---

Data Download Date/Time: 09/11/2019 00:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PYE,CHRISTOPHER,T  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: PYE,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 08/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: BHIMANI

DATE/TIME OF TEST: 08/10/2019 00:36:01

---

Diagnostics Check	OK	00:28
Air Blank	0.000	00:29
Control Test	0.080	00:29
Air Blank	0.000	00:29
Subject Sample # 1	0.000	00:30
Breath Volume	1.156	00:30
Air Blank	0.000	00:30
Air Blank	0.000	00:32
Subject Sample # 2	0.000	00:34
Breath Volume	2.218	00:34
Air Blank	0.000	00:35
Control Test	0.078	00:35
Air Blank	0.000	00:35
Diagnostics Check	OK	00:36

---

Data Download Date/Time: 09/11/2019 00:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PYE,CHRISTOPHER,T  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: MONTFORD,TODD  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 08/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 05:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:54

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 08/10/2019 05:28:51

---

Diagnostics Check	OK	05:22
Air Blank	0.000	05:23
Control Test	0.080	05:23
Air Blank	0.000	05:24
Subject Sample # 1	0.000	05:24
Breath Volume	3.300	05:24
Air Blank	0.000	05:25
Air Blank	0.000	05:27
Subject Sample # 2	0.000	05:27
Breath Volume	3.722	05:27
Air Blank	0.000	05:27
Control Test	0.078	05:28
Air Blank	0.000	05:28
Diagnostics Check	OK	05:28

---

Data Download Date/Time: 09/11/2019 00:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PYE,CHRISTOPHER,T  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: PYE,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 08/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: MATULA

DATE/TIME OF TEST: 08/12/2019 03:18:46

---

Diagnostics Check		OK	03:12
Air Blank		0.000	03:12
Control Test		0.079	03:13
Air Blank		0.000	03:13
Subject Sample	# 1	0.206	03:14
Breath Volume		2.355	03:14
Air Blank		0.000	03:14
Air Blank		0.000	03:16
Subject Sample	# 2	0.201	03:17
Breath Volume		1.835	03:17
Air Blank		0.000	03:17
Control Test		0.079	03:18
Air Blank		0.000	03:18
Diagnostics Check		OK	03:18

---

Data Download Date/Time: 09/11/2019 00:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRISWOLD,DAVID,J  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: DEOLEO,GEOVANNI  
ARRESTING OFFICER AGENCY: St. Cloud PD  
LAST AGENCY INSPECTION: 08/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: REED

DATE/TIME OF TEST: 08/13/2019 00:32:43

---

Diagnostics Check	OK	00:29	
Air Blank	0.000	00:30	
Control Test	0.079	00:30	
Air Blank	0.000	00:31	
Subject Sample	# 1	00:31	Subject Test Refused
Breath Volume	0.000	00:31	
Air Blank	0.000	00:31	
Control Test	0.079	00:32	
Air Blank	0.000	00:32	
Diagnostics Check	OK	00:32	

---

Data Download Date/Time: 09/11/2019 00:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LYONS,DANIEL,E  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: CRAWFORD,DAVID  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 08/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 08/14/2019 01:49:35

---

Diagnostics Check	OK	01:41
Air Blank	0.000	01:42
Control Test	0.080	01:42
Air Blank	0.000	01:43
Subject Sample # 1	0.125	01:43
Breath Volume	2.320	01:43
Air Blank	0.000	01:44
Air Blank	0.000	01:46
Subject Sample # 2	0.128	01:48
Breath Volume	1.898	01:48
Air Blank	0.000	01:48
Control Test	0.078	01:48
Air Blank	0.000	01:49
Diagnostics Check	OK	01:49

---

Data Download Date/Time: 09/11/2019 00:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FORD, GLENN, W  
BREATH TEST OPERATOR AGENCY: Kissimmee PD  
ARRESTING OFFICER: GAVILANES, MONICA  
ARRESTING OFFICER AGENCY: Kissimmee PD  
LAST AGENCY INSPECTION: 08/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 13:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:40

SUBJECT'S LAST NAME: POSSE

DATE/TIME OF TEST: 08/16/2019 14:07:54

---

Diagnostics Check	OK	14:01
Air Blank	0.000	14:01
Control Test	0.078	14:02
Air Blank	0.000	14:02
Subject Sample # 1	0.185	14:03
Breath Volume	1.683	14:03
Air Blank	0.000	14:04
Air Blank	0.000	14:05
Subject Sample # 2	0.194	14:06
Breath Volume	3.410	14:06
Air Blank	0.000	14:06
Control Test	0.077	14:07
Air Blank	0.000	14:07
Diagnostics Check	OK	14:07

---

Data Download Date/Time: 09/11/2019 00:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PYE,CHRISTOPHER,T  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: PYE,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 08/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: SMITH JR

DATE/TIME OF TEST: 08/20/2019 03:40:13

---

Diagnostics Check	OK	03:33
Air Blank	0.000	03:34
Control Test	0.079	03:34
Air Blank	0.000	03:34
Subject Sample # 1	0.131	03:35
Breath Volume	1.996	03:35
Air Blank	0.000	03:36
Air Blank	0.000	03:38
Subject Sample # 2	0.130	03:38
Breath Volume	1.683	03:38
Air Blank	0.000	03:39
Control Test	0.077	03:39
Air Blank	0.000	03:40
Diagnostics Check	OK	03:40

---

Data Download Date/Time: 09/11/2019 00:10



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PYE,CHRISTOPHER,T  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: REYES,ISAAC  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 08/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 08/25/2019 00:57:19

---

Diagnostics Check	OK	00:50
Air Blank	0.000	00:51
Control Test	0.079	00:51
Air Blank	0.000	00:52
Subject Sample # 1	0.113	00:52
Breath Volume	2.292	00:52
Air Blank	0.000	00:53
Air Blank	0.000	00:55
Subject Sample # 2	0.111	00:55
Breath Volume	2.011	00:55
Air Blank	0.000	00:56
Control Test	0.078	00:56
Air Blank	0.000	00:57
Diagnostics Check	OK	00:57

---

Data Download Date/Time: 09/11/2019 00:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LYONS,DANIEL,E  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: TORRES,EDGARDO  
ARRESTING OFFICER AGENCY: St. Cloud PD  
LAST AGENCY INSPECTION: 08/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 12:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: ROMAN SANABRIA

DATE/TIME OF TEST: 08/29/2019 02:08:41

---

Diagnostics Check	OK	02:01
Air Blank	0.000	02:02
Control Test	0.079	02:02
Air Blank	0.000	02:03
Subject Sample # 1	0.153	02:03
Breath Volume	1.621	02:03
Air Blank	0.000	02:04
Air Blank	0.000	02:06
Subject Sample # 2	0.151	02:07
Breath Volume	1.957	02:07
Air Blank	0.000	02:07
Control Test	0.078	02:08
Air Blank	0.000	02:08
Diagnostics Check	OK	02:08

---

Data Download Date/Time: 09/11/2019 00:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PYE,CHRISTOPHER,T  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: PYE,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 08/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:19

SUBJECT'S LAST NAME: GRACIA-RAMIREZ DE AR

DATE/TIME OF TEST: 08/29/2019 21:48:06

---

Diagnostics Check	OK	21:40
Air Blank	0.000	21:41
Control Test	0.080	21:41
Air Blank	0.000	21:42
Subject Sample # 1	0.055	21:43
Breath Volume	1.882	21:43
Air Blank	0.000	21:44
Air Blank	0.000	21:45
Subject Sample # 2	0.055	21:46
Breath Volume	1.558	21:46
Air Blank	0.000	21:47
Control Test	0.079	21:47
Air Blank	0.000	21:48
Diagnostics Check	OK	21:48

---

Data Download Date/Time: 09/11/2019 00:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PYE,CHRISTOPHER,T  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: CAMPOS,BARBARA  
ARRESTING OFFICER AGENCY: Kissimmee PD  
LAST AGENCY INSPECTION: 08/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: DE JESUS LUGO

DATE/TIME OF TEST: 08/30/2019 03:42:50

---

Diagnostics Check	OK	03:36
Air Blank	0.000	03:36
Control Test	0.078	03:36
Air Blank	0.000	03:37
Subject Sample # 1	0.143	03:38
Breath Volume	1.101	03:38
Air Blank	0.000	03:39
Air Blank	0.000	03:40
Subject Sample # 2	0.148	03:41
Breath Volume	1.234	03:41
Air Blank	0.000	03:41
Control Test	0.078	03:42
Air Blank	0.000	03:42
Diagnostics Check	OK	03:42

---

Data Download Date/Time: 09/11/2019 00:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BAEZ,BAYRON,L  
BREATH TEST OPERATOR AGENCY: Kissimmee PD  
ARRESTING OFFICER: SANTONE,LACEY  
ARRESTING OFFICER AGENCY: Kissimmee PD  
LAST AGENCY INSPECTION: 08/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: DIAZ LOPEZ

DATE/TIME OF TEST: 08/31/2019 03:52:52

---

Diagnostics Check	OK	03:46
Air Blank	0.000	03:47
Control Test	0.079	03:47
Air Blank	0.000	03:47
Subject Sample # 1	0.138	03:48
Breath Volume	3.582	03:48
Air Blank	0.000	03:49
Air Blank	0.000	03:50
Subject Sample # 2	0.131	03:51
Breath Volume	3.226	03:51
Air Blank	0.000	03:51
Control Test	0.079	03:52
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

---

Data Download Date/Time: 09/11/2019 00:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LYONS,DANIEL,E  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: OCHOA,YAMIANIS  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 08/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: URTENBERGER III

DATE/TIME OF TEST: 08/03/2019 02:16:48

---

Diagnostics Check	OK	02:10
Air Blank	0.000	02:10
Control Test	0.079	02:10
Air Blank	0.000	02:11
Subject Sample # 1	0.183	02:12
Breath Volume	2.093	02:12
Air Blank	0.000	02:12
Air Blank	0.000	02:14
Subject Sample # 2	0.181	02:15
Breath Volume	1.730	02:15
Air Blank	0.000	02:15
Control Test	0.078	02:16
Air Blank	0.000	02:16
Diagnostics Check	OK	02:16

---

Data Download Date/Time: 09/11/2019 00:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PAYNE,T,M  
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD  
ARRESTING OFFICER: BROWN,D  
ARRESTING OFFICER AGENCY: Cocoa Beach PD  
LAST AGENCY INSPECTION: 07/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A3  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:27

SUBJECT'S LAST NAME: SAUNDERS

DATE/TIME OF TEST: 08/09/2019 20:57:47

---

Diagnostics Check	OK	20:51
Air Blank	0.000	20:52
Control Test	0.078	20:52
Air Blank	0.000	20:52
Subject Sample # 1	0.228	20:53
Breath Volume	2.066	20:53
Air Blank	0.000	20:53
Air Blank	0.000	20:55
Subject Sample # 2	0.228	20:56
Breath Volume	1.824	20:56
Air Blank	0.000	20:56
Control Test	0.079	20:57
Air Blank	0.000	20:57
Diagnostics Check	OK	20:57

---

Data Download Date/Time: 09/06/2019 13:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS,J,  
BREATH TEST OPERATOR AGENCY: Titusville PD  
ARRESTING OFFICER: WILLIAMS,J  
ARRESTING OFFICER AGENCY: Titusville PD  
LAST AGENCY INSPECTION: 07/31/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29017080A2  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 19:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: MOSHER

DATE/TIME OF TEST: 08/04/2019 20:20:38

---

Diagnostics Check	OK	20:14
Air Blank	0.000	20:14
Control Test	0.081	20:15
Air Blank	0.000	20:15
Subject Sample # 1	0.133	20:16
Breath Volume	2.109	20:16
Air Blank	0.000	20:16
Air Blank	0.000	20:18
Subject Sample # 2	0.132	20:19
Breath Volume	1.582	20:19
Air Blank	0.000	20:19
Control Test	0.080	20:20
Air Blank	0.000	20:20
Diagnostics Check	OK	20:20

---

Data Download Date/Time: 09/03/2019 10:36



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HANSON,JASON,A  
BREATH TEST OPERATOR AGENCY: Titusville PD  
ARRESTING OFFICER: BOND,BRAIN  
ARRESTING OFFICER AGENCY: Titusville PD  
LAST AGENCY INSPECTION: 07/31/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29017080A2  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: WEITZEL

DATE/TIME OF TEST: 08/06/2019 02:33:29

---

Diagnostics Check		OK	02:26
Air Blank		0.000	02:26
Control Test		0.080	02:27
Air Blank		0.000	02:27
Subject Sample	# 1	0.125	02:28
Breath Volume		3.617	02:28
Air Blank		0.000	02:29
Air Blank		0.000	02:31
Subject Sample	# 2	0.123	02:31
Breath Volume		3.789	02:31
Air Blank		0.000	02:32
Control Test		0.079	02:32
Air Blank		0.000	02:33
Diagnostics Check		OK	02:33

---

Data Download Date/Time: 09/03/2019 10:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WEAVER,JACOB,T  
BREATH TEST OPERATOR AGENCY: Titusville PD  
ARRESTING OFFICER: BOND,B  
ARRESTING OFFICER AGENCY: Titusville PD  
LAST AGENCY INSPECTION: 07/31/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29017080A2  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:47

SUBJECT'S LAST NAME: RIDDLE

DATE/TIME OF TEST: 08/06/2019 22:30:51

---

Diagnostics Check	OK	22:24
Air Blank	0.000	22:24
Control Test	0.081	22:25
Air Blank	0.000	22:25
Subject Sample # 1	0.000	22:26
Breath Volume	3.398	22:26
Air Blank	0.000	22:26
Air Blank	0.000	22:28
Subject Sample # 2	0.000	22:29
Breath Volume	2.769	22:29
Air Blank	0.000	22:29
Control Test	0.080	22:30
Air Blank	0.000	22:30
Diagnostics Check	OK	22:30

---

Data Download Date/Time: 09/03/2019 10:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WEAVER,JACOB,T  
BREATH TEST OPERATOR AGENCY: Titusville PD  
ARRESTING OFFICER: BOND,BRIAN  
ARRESTING OFFICER AGENCY: Titusville PD  
LAST AGENCY INSPECTION: 07/31/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29017080A2  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:49  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: ST LOUIS

DATE/TIME OF TEST: 08/10/2019 03:52:34

---

Diagnostics Check	OK	03:43
Air Blank	0.000	03:43
Control Test	0.080	03:44
Air Blank	0.000	03:44
Subject Sample # 1	0.095	03:46
Breath Volume	1.675	03:46
Air Blank	0.000	03:47
Air Blank	0.000	03:49
Subject Sample # 2	0.091	03:51
Breath Volume	1.964	03:51
Air Blank	0.000	03:51
Control Test	0.079	03:51
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

---

Data Download Date/Time: 09/03/2019 10:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCDOLE,GEOFFERY,E  
BREATH TEST OPERATOR AGENCY: Titusville PD  
ARRESTING OFFICER: BLOUGOURAS,Z  
ARRESTING OFFICER AGENCY: Titusville PD  
LAST AGENCY INSPECTION: 07/31/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29017080A2  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 15:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 15:45

SUBJECT'S LAST NAME: BURNELL

DATE/TIME OF TEST: 08/28/2019 16:22:40

---

Diagnostics Check	OK	16:15
Air Blank	0.000	16:15
Control Test	0.081	16:16
Air Blank	0.000	16:16
Subject Sample # 1	0.030	16:17
Breath Volume	2.277	16:17
Air Blank	0.000	16:18
Air Blank	0.000	16:20
Subject Sample # 2	0.030	16:21
Breath Volume	1.804	16:21
Air Blank	0.000	16:21
Control Test	0.080	16:22
Air Blank	0.000	16:22
Diagnostics Check	OK	16:22

---

Data Download Date/Time: 09/03/2019 10:36