

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATSCH,JOSHUA,K
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: PATSCH,JOSHUA
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 08/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: MILLER JR

DATE/TIME OF TEST: 09/21/2019 23:35:06

Diagnostics Check	OK	23:28
Air Blank	0.000	23:29
Control Test	0.081	23:29
Air Blank	0.000	23:30
Subject Sample # 1	0.250	23:30
Breath Volume	2.542	23:30
Air Blank	0.000	23:31
Air Blank	0.000	23:33
Subject Sample # 2	0.232	23:33
Breath Volume	1.554	23:33
Air Blank	0.000	23:34
Control Test	0.079	23:34
Air Blank	0.000	23:35
Diagnostics Check	OK	23:35

Data Download Date/Time: 09/29/2019 13:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATSCH,JOSHUA,K
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: NAYLOR,MESHELLE
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 08/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: SENEFF

DATE/TIME OF TEST: 09/23/2019 03:34:34

Diagnostics Check	OK	03:27
Air Blank	0.000	03:28
Control Test	0.080	03:28
Air Blank	0.000	03:29
Subject Sample # 1	0.074	03:29
Breath Volume	3.617	03:29
Air Blank	0.000	03:30
Air Blank	0.000	03:32
Subject Sample # 2	0.057	03:33
Breath Volume	1.980	03:33
Air Blank	0.000	03:33
Control Test	0.080	03:33
Air Blank	0.000	03:34
Diagnostics Check	OK	03:34

Data Download Date/Time: 09/29/2019 13:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DA SILVA, ANDRE, F
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: QUADE, MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 05:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:45

SUBJECT'S LAST NAME: RAMOS BENITEZ

DATE/TIME OF TEST: 09/01/2019 06:13:22

Diagnostics Check	OK	06:06
Air Blank	0.000	06:06
Control Test	0.080	06:06
Air Blank	0.000	06:07
Subject Sample # 1	0.171	06:08
Breath Volume	3.195	06:08
Air Blank	0.000	06:09
Air Blank	0.000	06:11
Subject Sample # 2	0.152	06:11
Breath Volume	2.476	06:11
Air Blank	0.000	06:12
Control Test	0.079	06:12
Air Blank	0.000	06:13
Diagnostics Check	OK	06:13

Data Download Date/Time: 09/05/2019 14:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BALLOUN,SAMUEL,C
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: GIANOTTA,DOMINICK
ARRESTING OFFICER AGENCY: St. Augustine Beach
LAST AGENCY INSPECTION: 08/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: WALCK

DATE/TIME OF TEST: 09/01/2019 23:03:55

Diagnostics Check	OK	22:49	
Air Blank	0.000	22:49	
Control Test	0.080	22:49	
Air Blank	0.000	22:50	
Subject Sample	# 1 0.173	22:51	
Breath Volume	1.441	22:51	
Air Blank	0.000	22:51	
Air Blank	0.000	22:53	
Subject Sample	# 2 0.159	22:56	Slope Not Level
Breath Volume	1.683	22:56	
Air Blank	0.000	22:57	
Air Blank	0.000	22:59	
Subject Sample	# 3 0.154	23:02	Volume Not Met
Breath Volume	0.597	23:02	
Air Blank	0.000	23:02	
Control Test	0.078	23:03	
Air Blank	0.000	23:03	
Diagnostics Check	OK	23:03	

Data Download Date/Time: 09/05/2019 14:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,THOMAS,H
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: BRIGGS,JASON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: COLUCCI

DATE/TIME OF TEST: 09/09/2019 02:21:22

Diagnostics Check	OK	02:14
Air Blank	0.000	02:14
Control Test	0.080	02:14
Air Blank	0.000	02:15
Subject Sample # 1	0.192	02:16
Breath Volume	4.066	02:16
Air Blank	0.000	02:17
Air Blank	0.000	02:18
Subject Sample # 2	0.176	02:19
Breath Volume	3.824	02:19
Air Blank	0.000	02:20
Control Test	0.079	02:20
Air Blank	0.000	02:21
Diagnostics Check	OK	02:21

Data Download Date/Time: 10/30/2019 12:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: CAMDEN,DENNIS
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 09/09/2019 04:20:18

Diagnostics Check	OK	04:08	
Air Blank	0.000	04:09	
Control Test	0.079	04:09	
Air Blank	0.000	04:10	
Subject Sample	# 1 0.151	04:13	Volume Not Met
Breath Volume	0.054	04:13	
Air Blank	0.000	04:13	
Air Blank	0.000	04:15	
Subject Sample	# 2 0.141	04:18	Volume Not Met
Breath Volume	0.121	04:18	
Air Blank	0.000	04:19	
Control Test	0.079	04:19	
Air Blank	0.000	04:20	
Diagnostics Check	OK	04:20	

Data Download Date/Time: 10/30/2019 12:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIS, TRAVIS, W
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: HETZLER, BRANDON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 12:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 09/12/2019 01:36:59

Diagnostics Check	OK	01:27
Air Blank	0.000	01:28
Control Test	0.080	01:28
Air Blank	0.000	01:29
Subject Sample # 1	0.127	01:31
Breath Volume	1.328	01:31
Air Blank	0.000	01:32
Air Blank	0.000	01:34
Subject Sample # 2	0.123	01:35
Breath Volume	1.808	01:35
Air Blank	0.000	01:36
Control Test	0.079	01:36
Air Blank	0.000	01:36
Diagnostics Check	OK	01:36

Data Download Date/Time: 10/30/2019 12:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,SEAN,C
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: CLARKE,SALLY
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 14:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:36

SUBJECT'S LAST NAME: KOZARK

DATE/TIME OF TEST: 09/15/2019 15:57:33

Diagnostics Check	OK	15:56	
Air Blank	0.000	15:56	
Control Test	0.080	15:57	
Air Blank		15:57	RFI Detect
Air Blank		15:57	RFI Detect

Data Download Date/Time: 10/30/2019 12:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,SEAN,C
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: CLARKE,SALLY
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 14:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:38

SUBJECT'S LAST NAME: KOZAK

DATE/TIME OF TEST: 09/15/2019 16:06:47

Diagnostics Check	OK	15:59
Air Blank	0.000	16:00
Control Test	0.079	16:00
Air Blank	0.000	16:01
Subject Sample # 1	0.147	16:02
Breath Volume	1.933	16:02
Air Blank	0.000	16:02
Air Blank	0.000	16:04
Subject Sample # 2	0.142	16:05
Breath Volume	1.816	16:05
Air Blank	0.000	16:05
Control Test	0.079	16:06
Air Blank	0.000	16:06
Diagnostics Check	OK	16:06

Data Download Date/Time: 10/30/2019 12:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASH,JUSTIN,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,DONAVAN
ARRESTING OFFICER AGENCY: St. Augustine Beach
LAST AGENCY INSPECTION: 09/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:23

SUBJECT'S LAST NAME: MASTERS

DATE/TIME OF TEST: 09/16/2019 22:56:17

Diagnostics Check	OK	22:46	
Air Blank	0.000	22:46	
Control Test	0.080	22:46	
Air Blank	0.000	22:47	
Subject Sample	# 1 0.184	22:50	Volume Not Met
Breath Volume	0.687	22:50	
Air Blank	0.000	22:51	
Air Blank	0.000	22:52	
Subject Sample	# 2	22:54	Subject Test Refused
Breath Volume	0.000	22:54	
Air Blank	0.000	22:55	
Control Test	0.078	22:55	
Air Blank	0.000	22:56	
Diagnostics Check	OK	22:56	

Data Download Date/Time: 10/30/2019 12:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASH,JUSTIN,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: CASH,JUSTIN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: ZINK

DATE/TIME OF TEST: 09/17/2019 03:14:08

Diagnostics Check	OK	03:08	
Air Blank	0.000	03:09	
Control Test	0.079	03:09	
Air Blank	0.000	03:10	
Subject Sample	# 1	03:12	Subject Test Refused
Breath Volume	0.000	03:12	
Air Blank	0.000	03:13	
Control Test	0.080	03:13	
Air Blank	0.000	03:14	
Diagnostics Check	OK	03:14	

Data Download Date/Time: 10/30/2019 12:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASH,JUSTIN,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: BRIGGS,JASON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: HANCOCK

DATE/TIME OF TEST: 09/17/2019 03:27:00

Diagnostics Check	OK	03:23	
Air Blank	0.000	03:24	
Control Test	0.080	03:24	
Air Blank	0.000	03:25	
Subject Sample	# 1	03:25	Subject Test Refused
Breath Volume	0.000	03:25	
Air Blank	0.000	03:26	
Control Test	0.080	03:26	
Air Blank	0.000	03:26	
Diagnostics Check	OK	03:27	

Data Download Date/Time: 10/30/2019 12:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEORGE,KYLE,J
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GIANNOTTA,D
ARRESTING OFFICER AGENCY: St. Augustine Beach
LAST AGENCY INSPECTION: 09/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 15:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:57

SUBJECT'S LAST NAME: WARD

DATE/TIME OF TEST: 09/17/2019 16:27:06

Diagnostics Check	OK	16:20
Air Blank	0.000	16:21
Control Test	0.079	16:21
Air Blank	0.000	16:22
Subject Sample # 1	0.165	16:22
Breath Volume	2.355	16:22
Air Blank	0.000	16:23
Air Blank	0.000	16:25
Subject Sample # 2	0.162	16:25
Breath Volume	2.460	16:25
Air Blank	0.000	16:26
Control Test	0.079	16:26
Air Blank	0.000	16:27
Diagnostics Check	OK	16:27

Data Download Date/Time: 10/30/2019 12:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: BLESSING

DATE/TIME OF TEST: 09/18/2019 02:13:05

Diagnostics Check	OK	02:07
Air Blank	0.000	02:07
Control Test	0.080	02:07
Air Blank	0.000	02:08
Subject Sample # 1	0.193	02:08
Breath Volume	2.046	02:08
Air Blank	0.000	02:09
Air Blank	0.000	02:11
Subject Sample # 2	0.199	02:11
Breath Volume	2.601	02:11
Air Blank	0.000	02:12
Control Test	0.080	02:12
Air Blank	0.000	02:13
Diagnostics Check	OK	02:13

Data Download Date/Time: 10/30/2019 12:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEORGE,KYLE,J
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: BROWN,B
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 16:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:39

SUBJECT'S LAST NAME: TENA

DATE/TIME OF TEST: 09/21/2019 17:05:02

Diagnostics Check	OK	17:00	
Air Blank	0.000	17:01	
Control Test	0.080	17:01	
Air Blank	0.000	17:02	
Subject Sample	# 1	17:03	Subject Test Refused
Breath Volume	0.000	17:03	
Air Blank	0.000	17:04	
Control Test	0.080	17:04	
Air Blank	0.000	17:04	
Diagnostics Check	OK	17:05	

Data Download Date/Time: 10/30/2019 12:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,MASON,E
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: THOMPSON,MASON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 19:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:44

SUBJECT'S LAST NAME: ALFORD

DATE/TIME OF TEST: 09/22/2019 20:27:36

Diagnostics Check	OK	20:21
Air Blank	0.000	20:21
Control Test	0.080	20:21
Air Blank	0.000	20:22
Subject Sample # 1	0.216	20:23
Breath Volume	1.437	20:23
Air Blank	0.000	20:23
Air Blank	0.000	20:25
Subject Sample # 2	0.213	20:26
Breath Volume	1.417	20:26
Air Blank	0.000	20:26
Control Test	0.079	20:26
Air Blank	0.000	20:27
Diagnostics Check	OK	20:27

Data Download Date/Time: 10/30/2019 12:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRESTON,MIKAYLA,D
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: BRIGGS,JASON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 09/22/2019 23:11:44

Diagnostics Check	OK	23:08	
Air Blank	0.000	23:09	
Control Test	0.080	23:09	
Air Blank	0.000	23:10	
Subject Sample	# 1	23:10	Subject Test Refused
Breath Volume	0.000	23:10	
Air Blank	0.000	23:10	
Control Test	0.080	23:11	
Air Blank	0.000	23:11	
Diagnostics Check	OK	23:11	

Data Download Date/Time: 10/30/2019 12:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,A
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: WELLS

DATE/TIME OF TEST: 09/23/2019 00:13:34

Diagnostics Check	OK	23:58	
Air Blank	0.000	23:59	
Control Test	0.080	23:59	
Air Blank	0.000	00:00	
Subject Sample	# 1 0.082	00:00	
Breath Volume	2.203	00:00	
Air Blank	0.000	00:01	
Air Blank	0.000	00:03	
Subject Sample	# 2 0.071	00:06	Volume Not Met
Breath Volume	0.246	00:06	
Air Blank	0.000	00:07	
Air Blank	0.000	00:08	
Subject Sample	# 3 0.044	00:12	Volume Not Met
Breath Volume	0.617	00:12	
Air Blank	0.000	00:12	
Control Test	0.080	00:12	
Air Blank	0.000	00:13	
Diagnostics Check	OK	00:13	

Data Download Date/Time: 10/30/2019 12:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRESTON,MIKAYLA,D
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: BRIGGS,JASON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 09/23/2019 03:43:57

Diagnostics Check	OK	03:41	
Air Blank	0.000	03:41	
Control Test	0.080	03:41	
Air Blank	0.000	03:42	
Subject Sample	# 1	03:42	Subject Test Refused
Breath Volume	0.000	03:42	
Air Blank	0.000	03:43	
Control Test	0.080	03:43	
Air Blank	0.000	03:43	
Diagnostics Check	OK	03:43	

Data Download Date/Time: 10/30/2019 12:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRERA,ALEXANDER,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: CAMDEN,D
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 03:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: AARON

DATE/TIME OF TEST: 09/23/2019 05:05:20

Diagnostics Check	OK	04:57
Air Blank	0.000	04:58
Control Test	0.079	04:58
Air Blank	0.000	04:59
Subject Sample # 1	0.065	04:59
Breath Volume	2.117	04:59
Air Blank	0.000	05:00
Air Blank	0.000	05:02
Subject Sample # 2	0.065	05:03
Breath Volume	2.726	05:03
Air Blank	0.000	05:04
Control Test	0.080	05:04
Air Blank	0.000	05:05
Diagnostics Check	OK	05:05

Data Download Date/Time: 10/30/2019 12:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRESTON,MIKAYLA,D
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: TRUSCIO,JESSE
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:22

SUBJECT'S LAST NAME: PRODROMIDES

DATE/TIME OF TEST: 09/25/2019 21:02:13

Diagnostics Check	OK	20:59	
Air Blank	0.000	20:59	
Control Test	0.080	21:00	
Air Blank	0.000	21:00	
Subject Sample	# 1	21:00	Subject Test Refused
Breath Volume	0.000	21:00	
Air Blank	0.000	21:01	
Control Test	0.080	21:01	
Air Blank	0.000	21:02	
Diagnostics Check	OK	21:02	

Data Download Date/Time: 10/30/2019 12:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: FORESTO

DATE/TIME OF TEST: 09/01/2019 00:57:09

Diagnostics Check	OK	00:51
Air Blank	0.000	00:51
Control Test	0.078	00:51
Air Blank	0.000	00:52
Subject Sample # 1	0.000	00:52
Breath Volume	1.437	00:52
Air Blank	0.000	00:53
Air Blank	0.000	00:55
Subject Sample # 2	0.000	00:55
Breath Volume	1.207	00:55
Air Blank	0.000	00:56
Control Test	0.078	00:56
Air Blank	0.000	00:57
Diagnostics Check	OK	00:57

Data Download Date/Time: 09/05/2019 14:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BALLOUN,SAMUEL,C
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: GIANOTTA,DOMINICK
ARRESTING OFFICER AGENCY: St. Augustine Beach
LAST AGENCY INSPECTION: 08/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: WALCK

DATE/TIME OF TEST: 09/01/2019 23:19:18

Diagnostics Check	OK	23:09
Air Blank	0.000	23:09
Control Test	0.078	23:09
Air Blank	0.000	23:10
Subject Sample # 1	0.168	23:12
Breath Volume	1.535	23:12
Air Blank	0.000	23:12
Air Blank	0.000	23:14
Subject Sample # 2	0.173	23:17
Breath Volume	2.097	23:17
Air Blank	0.000	23:18
Control Test	0.080	23:18
Air Blank	0.000	23:19
Diagnostics Check	OK	23:19

Data Download Date/Time: 09/05/2019 14:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSARD,JOSHUA,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: BODKIN,MICHAEL
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: MEJIAS

DATE/TIME OF TEST: 09/01/2019 02:46:59

Diagnostics Check	OK	02:41	
Air Blank	0.000	02:41	
Control Test	0.080	02:41	
Air Blank	0.000	02:42	
Subject Sample	# 1	02:45	Subject Test Refused
Breath Volume	0.070	02:45	
Air Blank	0.000	02:46	
Control Test	0.081	02:46	
Air Blank	0.000	02:46	
Diagnostics Check	OK	02:46	

Data Download Date/Time: 09/25/2019 19:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSARD,JOSHUA,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: MACDOWEL,MAYCAN
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: BARRETT

DATE/TIME OF TEST: 09/01/2019 03:21:52

Diagnostics Check	OK	03:15
Air Blank	0.000	03:16
Control Test	0.080	03:16
Air Blank	0.000	03:16
Subject Sample # 1	0.134	03:17
Breath Volume	1.875	03:17
Air Blank	0.000	03:18
Air Blank	0.000	03:19
Subject Sample # 2	0.132	03:20
Breath Volume	1.964	03:20
Air Blank	0.000	03:20
Control Test	0.080	03:21
Air Blank	0.000	03:21
Diagnostics Check	OK	03:21

Data Download Date/Time: 09/25/2019 19:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,RICHARD,L
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: GLASS,ALEC
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: BYRD

DATE/TIME OF TEST: 09/03/2019 05:15:22

Diagnostics Check	OK	05:05
Air Blank	0.000	05:05
Control Test	0.081	05:06
Air Blank	0.000	05:06
Subject Sample # 1	0.122	05:08
Breath Volume	2.511	05:08
Air Blank	0.000	05:09
Air Blank	0.000	05:11
Subject Sample # 2	0.117	05:13
Breath Volume	1.632	05:13
Air Blank	0.000	05:14
Control Test	0.079	05:14
Air Blank	0.000	05:15
Diagnostics Check	OK	05:15

Data Download Date/Time: 09/25/2019 19:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARO,AARON,M
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: MARTIN,STERLING
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 20:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:18

SUBJECT'S LAST NAME: HANSEN

DATE/TIME OF TEST: 09/07/2019 21:53:52

Diagnostics Check	OK	21:49	
Air Blank	0.000	21:50	
Control Test	0.081	21:50	
Air Blank	0.000	21:51	
Subject Sample	# 1	21:52	Subject Test Refused
Breath Volume	0.000	21:52	
Air Blank	0.000	21:52	
Control Test	0.081	21:53	
Air Blank	0.000	21:53	
Diagnostics Check	OK	21:53	

Data Download Date/Time: 09/25/2019 19:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNNE,JUSTIN,R
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: DUNNE,JUSTIN
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: WARREN

DATE/TIME OF TEST: 09/14/2019 02:54:17

Diagnostics Check	OK	02:47
Air Blank	0.000	02:47
Control Test	0.082	02:47
Air Blank	0.000	02:48
Subject Sample # 1	0.134	02:49
Breath Volume	1.441	02:49
Air Blank	0.000	02:50
Air Blank	0.000	02:52
Subject Sample # 2	0.137	02:52
Breath Volume	1.222	02:52
Air Blank	0.000	02:53
Control Test	0.079	02:53
Air Blank	0.000	02:54
Diagnostics Check	OK	02:54

Data Download Date/Time: 09/25/2019 19:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARO,AARON,M
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: BAZE,KAYLOR
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: FISHEL

DATE/TIME OF TEST: 09/21/2019 03:25:26

Diagnostics Check	OK	03:19
Air Blank	0.000	03:19
Control Test	0.081	03:19
Air Blank	0.000	03:20
Subject Sample # 1	0.160	03:21
Breath Volume	1.285	03:21
Air Blank	0.000	03:21
Air Blank	0.000	03:23
Subject Sample # 2	0.163	03:23
Breath Volume	1.589	03:23
Air Blank	0.000	03:24
Control Test	0.079	03:24
Air Blank	0.000	03:25
Diagnostics Check	OK	03:25

Data Download Date/Time: 09/25/2019 19:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNNE,JUSTIN,R
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: DUNNE,JUSTIN
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 08/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 22:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:23

SUBJECT'S LAST NAME: LALOMIA

DATE/TIME OF TEST: 09/23/2019 22:50:13

Diagnostics Check	OK	22:43
Air Blank	0.000	22:43
Control Test	0.081	22:44
Air Blank	0.000	22:44
Subject Sample # 1	0.000	22:45
Breath Volume	1.316	22:45
Air Blank	0.000	22:45
Air Blank	0.000	22:47
Subject Sample # 2	0.000	22:48
Breath Volume	1.394	22:48
Air Blank	0.000	22:49
Control Test	0.080	22:49
Air Blank	0.000	22:50
Diagnostics Check	OK	22:50

Data Download Date/Time: 09/25/2019 19:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,STERLING,R
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: MARTIN,STERLING
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 09/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 21:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: BURGDORF

DATE/TIME OF TEST: 09/25/2019 22:19:31

Diagnostics Check	OK	22:10
Air Blank	0.000	22:11
Control Test	0.081	22:11
Air Blank	0.000	22:12
Subject Sample # 1	0.260	22:13
Breath Volume	2.515	22:13
Air Blank	0.000	22:14
Air Blank	0.000	22:16
Subject Sample # 2	0.257	22:17
Breath Volume	2.769	22:17
Air Blank	0.000	22:18
Control Test	0.080	22:18
Air Blank	0.000	22:19
Diagnostics Check	OK	22:19

Data Download Date/Time: 10/28/2019 15:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TWING,BRANDON,R
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: SCHERR,E
ARRESTING OFFICER AGENCY: Flagler Beach PD
LAST AGENCY INSPECTION: 08/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: D'ARCY

DATE/TIME OF TEST: 09/01/2019 00:54:01

Diagnostics Check	OK	00:47
Air Blank	0.000	00:48
Control Test	0.081	00:48
Air Blank	0.000	00:48
Subject Sample # 1	0.116	00:49
Breath Volume	1.468	00:49
Air Blank	0.000	00:50
Air Blank	0.000	00:52
Subject Sample # 2	0.112	00:52
Breath Volume	1.539	00:52
Air Blank	0.000	00:53
Control Test	0.080	00:53
Air Blank	0.000	00:53
Diagnostics Check	OK	00:54

Data Download Date/Time: 09/05/2019 00:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIMA,MICHAEL,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: LIMA,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: RAINEY

DATE/TIME OF TEST: 09/01/2019 01:06:26

Diagnostics Check	OK	00:59
Air Blank	0.000	01:00
Control Test	0.080	01:00
Air Blank	0.000	01:00
Subject Sample # 1	0.126	01:02
Breath Volume	1.535	01:02
Air Blank	0.000	01:02
Air Blank	0.000	01:04
Subject Sample # 2	0.137	01:04
Breath Volume	2.109	01:04
Air Blank	0.000	01:05
Control Test	0.080	01:05
Air Blank	0.000	01:06
Diagnostics Check	OK	01:06

Data Download Date/Time: 09/05/2019 00:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TWING,BRANDON,R
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: CONWAY,P
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 08/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 09/02/2019 04:15:11

Diagnostics Check	OK	04:09
Air Blank	0.000	04:09
Control Test	0.082	04:09
Air Blank	0.000	04:10
Subject Sample # 1	0.156	04:10
Breath Volume	2.359	04:10
Air Blank	0.000	04:11
Air Blank	0.000	04:13
Subject Sample # 2	0.154	04:13
Breath Volume	1.429	04:13
Air Blank	0.000	04:14
Control Test	0.080	04:14
Air Blank	0.000	04:15
Diagnostics Check	OK	04:15

Data Download Date/Time: 09/05/2019 00:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TWING,BRANDON,R
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: GADDIE,K
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 08/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: SPARLING

DATE/TIME OF TEST: 09/10/2019 01:37:32

Diagnostics Check	OK	01:29	
Air Blank	0.000	01:29	
Control Test	0.080	01:29	
Air Blank	0.000	01:30	
Subject Sample	# 1 0.085	01:33	Volume Not Met
Breath Volume	0.167	01:33	
Air Blank	0.000	01:34	
Air Blank	0.000	01:35	
Subject Sample	# 2	01:36	Subject Test Refused
Breath Volume	0.000	01:36	
Air Blank	0.000	01:36	
Control Test	0.079	01:36	
Air Blank	0.000	01:37	
Diagnostics Check	OK	01:37	

Data Download Date/Time: 09/28/2019 15:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NICOLAS,MITA,
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: MYERS,ROBERT
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 08/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: LUCKETT

DATE/TIME OF TEST: 09/14/2019 01:38:19

Diagnostics Check	OK	01:31
Air Blank	0.000	01:31
Control Test	0.081	01:31
Air Blank	0.000	01:32
Subject Sample # 1	0.155	01:33
Breath Volume	1.816	01:33
Air Blank	0.000	01:34
Air Blank	0.000	01:36
Subject Sample # 2	0.140	01:36
Breath Volume	1.703	01:36
Air Blank	0.000	01:37
Control Test	0.079	01:37
Air Blank	0.000	01:38
Diagnostics Check	OK	01:38

Data Download Date/Time: 09/28/2019 15:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARPER,DARIN,W
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HARPER,DARIN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: MEADOWS

DATE/TIME OF TEST: 09/15/2019 04:01:35

Diagnostics Check	OK	03:54
Air Blank	0.000	03:54
Control Test	0.080	03:54
Air Blank	0.000	03:55
Subject Sample # 1	0.133	03:56
Breath Volume	1.789	03:56
Air Blank	0.000	03:56
Air Blank	0.000	03:58
Subject Sample # 2	0.130	04:00
Breath Volume	2.707	04:00
Air Blank	0.000	04:00
Control Test	0.079	04:01
Air Blank	0.000	04:01
Diagnostics Check	OK	04:01

Data Download Date/Time: 09/28/2019 15:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NICOLAS,MITA,
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: CONWAY,PHILIP
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 08/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: KIRUTHI

DATE/TIME OF TEST: 09/16/2019 00:37:45

Diagnostics Check	OK	00:30
Air Blank	0.000	00:30
Control Test	0.080	00:30
Air Blank	0.000	00:31
Subject Sample # 1	0.179	00:32
Breath Volume	1.128	00:32
Air Blank	0.000	00:33
Air Blank	0.000	00:35
Subject Sample # 2	0.177	00:36
Breath Volume	1.406	00:36
Air Blank	0.000	00:36
Control Test	0.079	00:37
Air Blank	0.000	00:37
Diagnostics Check	OK	00:37

Data Download Date/Time: 09/28/2019 15:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIMA,MICHAEL,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: LIMA,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: ROBERSON

DATE/TIME OF TEST: 09/16/2019 04:03:33

Diagnostics Check	OK	03:57
Air Blank	0.000	03:57
Control Test	0.080	03:57
Air Blank	0.000	03:58
Subject Sample # 1	0.191	03:59
Breath Volume	2.000	03:59
Air Blank	0.000	03:59
Air Blank	0.000	04:01
Subject Sample # 2	0.201	04:02
Breath Volume	2.171	04:02
Air Blank	0.000	04:02
Control Test	0.080	04:02
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time: 09/28/2019 15:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY, WILLIAM, J
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: BOURKE, JOSEPH
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 08/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: SPAULDING

DATE/TIME OF TEST: 09/18/2019 00:04:13

Diagnostics Check	OK	00:01	
Air Blank	0.000	00:01	
Control Test	0.081	00:02	
Air Blank	0.000	00:02	
Subject Sample	# 1	00:02	Subject Test Refused
Breath Volume	0.000	00:02	
Air Blank	0.000	00:03	
Control Test	0.081	00:03	
Air Blank	0.000	00:04	
Diagnostics Check	OK	00:04	

Data Download Date/Time: 09/28/2019 15:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TWING,BRANDON,R
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: WOODCOCK,M
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 08/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: MCCUE

DATE/TIME OF TEST: 09/20/2019 05:23:30

Diagnostics Check	OK	05:20	
Air Blank	0.000	05:21	
Control Test	0.081	05:21	
Air Blank	0.000	05:21	
Subject Sample	# 1	05:22	Subject Test Refused
Breath Volume	0.000	05:22	
Air Blank	0.000	05:22	
Control Test	0.079	05:22	
Air Blank	0.000	05:23	
Diagnostics Check	OK	05:23	

Data Download Date/Time: 09/28/2019 15:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,CARL,L
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: PARKER,CARL
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 08/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 09/22/2019 03:34:01

Diagnostics Check	OK	03:27
Air Blank	0.000	03:28
Control Test	0.080	03:28
Air Blank	0.000	03:28
Subject Sample # 1	0.161	03:29
Breath Volume	3.519	03:29
Air Blank	0.000	03:30
Air Blank	0.000	03:32
Subject Sample # 2	0.150	03:32
Breath Volume	2.507	03:32
Air Blank	0.000	03:33
Control Test	0.080	03:33
Air Blank	0.000	03:33
Diagnostics Check	OK	03:34

Data Download Date/Time: 09/28/2019 15:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,CARL,L
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: PARKER,CARL
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 08/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: LEFEBVRE

DATE/TIME OF TEST: 09/23/2019 04:22:12

Diagnostics Check	OK	04:19	
Air Blank	0.000	04:19	
Control Test	0.079	04:20	
Air Blank	0.000	04:20	
Subject Sample	# 1	04:20	Subject Test Refused
Breath Volume	0.000	04:20	
Air Blank	0.000	04:21	
Control Test	0.079	04:21	
Air Blank	0.000	04:22	
Diagnostics Check	OK	04:22	

Data Download Date/Time: 09/28/2019 15:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY, WILLIAM, J
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: LOGSDON, DEREK
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 08/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: FERREIRA

DATE/TIME OF TEST: 09/26/2019 04:07:31

Diagnostics Check	OK	04:00
Air Blank	0.000	04:01
Control Test	0.080	04:01
Air Blank	0.000	04:02
Subject Sample # 1	0.259	04:02
Breath Volume	1.593	04:02
Air Blank	0.000	04:03
Air Blank	0.000	04:05
Subject Sample # 2	0.254	04:05
Breath Volume	1.398	04:05
Air Blank	0.000	04:06
Control Test	0.080	04:06
Air Blank	0.000	04:07
Diagnostics Check	OK	04:07

Data Download Date/Time: 09/28/2019 15:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCARTHY, EVAN, V
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: MCCARTHY, EVAN
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 08/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 07:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:05

SUBJECT'S LAST NAME: CHASSE

DATE/TIME OF TEST: 09/26/2019 08:36:27

Diagnostics Check	OK	08:30
Air Blank	0.000	08:30
Control Test	0.079	08:30
Air Blank	0.000	08:31
Subject Sample # 1	0.401	08:31
Breath Volume	2.066	08:31
Air Blank	0.000	08:32
Air Blank	0.000	08:34
Subject Sample # 2	0.388	08:34
Breath Volume	3.921	08:34
Air Blank	0.000	08:35
Control Test	0.080	08:35
Air Blank	0.000	08:36
Diagnostics Check	OK	08:36

Data Download Date/Time: 09/28/2019 15:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONG,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: ROSSI,CRAIG
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 08/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:10

SUBJECT'S LAST NAME: SHUMATE

DATE/TIME OF TEST: 09/26/2019 16:35:55

Diagnostics Check	OK	16:32	
Air Blank	0.000	16:33	
Control Test	0.080	16:33	
Air Blank	0.000	16:33	
Subject Sample	# 1	16:34	Subject Test Refused
Breath Volume	0.000	16:34	
Air Blank	0.000	16:35	
Control Test	0.080	16:35	
Air Blank	0.000	16:35	
Diagnostics Check	OK	16:35	

Data Download Date/Time: 09/28/2019 15:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: MANATEE COUNTY SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: MANATEE COUNTY SO
LAST AGENCY INSPECTION: 09/17/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 18:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:13

SUBJECT'S LAST NAME: FRANCISCO DIEGO

DATE/TIME OF TEST: 09/19/2019 18:40:14

Diagnostics Check	OK	18:33
Air Blank	0.000	18:34
Control Test	0.081	18:34
Air Blank	0.000	18:34
Subject Sample # 1	0.271	18:35
Breath Volume	3.292	18:35
Air Blank	0.000	18:36
Air Blank	0.000	18:37
Subject Sample # 2	0.260	18:38
Breath Volume	3.140	18:38
Air Blank	0.000	18:39
Control Test	0.079	18:39
Air Blank	0.000	18:40
Diagnostics Check	OK	18:40

Data Download Date/Time: 10/21/2019 09:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: MANATEE COUNTY SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: MANATEE COUNTY SO
LAST AGENCY INSPECTION: 09/17/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 21:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: MUGGRIDGE

DATE/TIME OF TEST: 09/20/2019 22:52:23

Diagnostics Check	OK	22:49	
Air Blank	0.000	22:49	
Control Test	0.081	22:50	
Air Blank	0.000	22:50	
Subject Sample	# 1	22:50	Subject Test Refused
Breath Volume	0.000	22:50	
Air Blank	0.000	22:51	
Control Test	0.081	22:51	
Air Blank	0.000	22:52	
Diagnostics Check	OK	22:52	

Data Download Date/Time: 10/21/2019 09:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: MANATEE COUNTY SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: MANATEE COUNTY SO
LAST AGENCY INSPECTION: 09/17/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 19:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: DOUTHITT

DATE/TIME OF TEST: 09/22/2019 20:16:50

Diagnostics Check	OK	20:10
Air Blank	0.000	20:11
Control Test	0.081	20:11
Air Blank	0.000	20:11
Subject Sample # 1	0.000	20:12
Breath Volume	2.359	20:12
Air Blank	0.000	20:12
Air Blank	0.000	20:14
Subject Sample # 2	0.000	20:15
Breath Volume	2.386	20:15
Air Blank	0.000	20:15
Control Test	0.081	20:16
Air Blank	0.000	20:16
Diagnostics Check	OK	20:16

Data Download Date/Time: 10/21/2019 09:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: MANATEE COUNTY SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: MANATEE COUNTY SO
LAST AGENCY INSPECTION: 09/17/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: RAMIREZ CINTO

DATE/TIME OF TEST: 09/26/2019 02:38:10

Diagnostics Check	OK	02:31
Air Blank	0.000	02:31
Control Test	0.081	02:32
Air Blank	0.000	02:32
Subject Sample # 1	0.195	02:33
Breath Volume	2.484	02:33
Air Blank	0.000	02:34
Air Blank	0.000	02:35
Subject Sample # 2	0.189	02:36
Breath Volume	3.347	02:36
Air Blank	0.000	02:37
Control Test	0.080	02:37
Air Blank	0.000	02:38
Diagnostics Check	OK	02:38

Data Download Date/Time: 10/21/2019 09:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KALLMEYER,DANIEL,K
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: N/A,N/A
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 08/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: PROB/PAROL
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: KOEHLER

DATE/TIME OF TEST: 09/04/2019 09:46:02

Diagnostics Check	OK	09:40
Air Blank	0.000	09:40
Control Test	0.081	09:40
Air Blank	0.000	09:41
Subject Sample # 1	0.000	09:41
Breath Volume	3.949	09:41
Air Blank	0.000	09:42
Air Blank	0.000	09:44
Subject Sample # 2	0.000	09:44
Breath Volume	4.222	09:44
Air Blank	0.000	09:45
Control Test	0.081	09:45
Air Blank	0.000	09:45
Diagnostics Check	OK	09:46

Data Download Date/Time: 09/24/2019 09:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KALLMEYER,DANIEL,K
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: N/A,N/A
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 08/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: PROB/PAROL
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:45

SUBJECT'S LAST NAME: DEEMER

DATE/TIME OF TEST: 09/06/2019 09:22:33

Diagnostics Check	OK	09:15
Air Blank	0.000	09:16
Control Test	0.082	09:16
Air Blank	0.000	09:17
Subject Sample # 1	0.000	09:18
Breath Volume	1.578	09:18
Air Blank	0.000	09:18
Air Blank	0.000	09:20
Subject Sample # 2	0.000	09:21
Breath Volume	1.699	09:21
Air Blank	0.000	09:21
Control Test	0.082	09:21
Air Blank	0.000	09:22
Diagnostics Check	OK	09:22

Data Download Date/Time: 09/24/2019 09:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KALLMEYER,DANIEL,K
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: N/A,N/A
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 08/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: PROB/PAROL
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:00

SUBJECT'S LAST NAME: LAUX

DATE/TIME OF TEST: 09/09/2019 10:44:38

Diagnostics Check	OK	10:38
Air Blank	0.000	10:38
Control Test	0.082	10:39
Air Blank	0.000	10:39
Subject Sample # 1	0.000	10:40
Breath Volume	2.136	10:40
Air Blank	0.000	10:40
Air Blank	0.000	10:42
Subject Sample # 2	0.000	10:43
Breath Volume	2.609	10:43
Air Blank	0.000	10:43
Control Test	0.083	10:44
Air Blank	0.000	10:44
Diagnostics Check	OK	10:44

Data Download Date/Time: 09/24/2019 09:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARING,KELLY,A
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: WARING,KELLY
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 08/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: PROB/PAROL
ARREST TIME: 13:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:48

SUBJECT'S LAST NAME: ROMAN ROSA

DATE/TIME OF TEST: 09/09/2019 14:14:45

Diagnostics Check	OK	14:08
Air Blank	0.000	14:09
Control Test	0.082	14:09
Air Blank	0.000	14:10
Subject Sample # 1	0.000	14:10
Breath Volume	2.871	14:10
Air Blank	0.000	14:11
Air Blank	0.000	14:13
Subject Sample # 2	0.000	14:13
Breath Volume	2.671	14:13
Air Blank	0.000	14:13
Control Test	0.081	14:14
Air Blank	0.000	14:14
Diagnostics Check	OK	14:14

Data Download Date/Time: 09/24/2019 09:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARING,KELLY,A
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: WARING,KELLY
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 08/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: PROB/PAROL
ARREST TIME: 14:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: DEEMER

DATE/TIME OF TEST: 09/09/2019 14:30:31

Diagnostics Check	OK	14:24	
Air Blank	0.000	14:24	
Control Test	0.082	14:25	
Air Blank	0.000	14:25	
Subject Sample # 1	0.000	14:26	
Breath Volume	2.164	14:26	
Air Blank	0.000	14:26	
Air Blank	0.000	14:28	
Subject Sample # 2	0.000	14:28	
Breath Volume	1.210	14:28	
Air Blank	0.000	14:29	
Control Test	0.081	14:29	
Air Blank		14:30	RFI Detect
Air Blank	0.000	14:30	

Data Download Date/Time: 09/24/2019 09:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,STEVEN,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: EVANS,STEVEN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 08/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 04:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: CULBERTSON

DATE/TIME OF TEST: 09/19/2019 05:16:44

Diagnostics Check	OK	05:10
Air Blank	0.000	05:10
Control Test	0.082	05:11
Air Blank	0.000	05:11
Subject Sample # 1	0.102	05:12
Breath Volume	1.417	05:12
Air Blank	0.000	05:12
Air Blank	0.000	05:14
Subject Sample # 2	0.103	05:15
Breath Volume	1.734	05:15
Air Blank	0.000	05:15
Control Test	0.082	05:16
Air Blank	0.000	05:16
Diagnostics Check	OK	05:16

Data Download Date/Time: 09/24/2019 09:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUABA,JASON,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GUABA,JASON
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: INGRAM

DATE/TIME OF TEST: 09/23/2019 01:51:20

Diagnostics Check	OK	01:48	
Air Blank	0.000	01:48	
Control Test	0.082	01:49	
Air Blank	0.000	01:49	
Subject Sample	# 1	01:49	Subject Test Refused
Breath Volume	0.000	01:49	
Air Blank	0.000	01:50	
Control Test	0.083	01:50	
Air Blank	0.000	01:51	
Diagnostics Check	OK	01:51	

Data Download Date/Time: 09/24/2019 09:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001152

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HELMS,WILLIAM,A
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: HELMS,WILLIAM
ARRESTING OFFICER AGENCY: New Smyrna Beach PD
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A5
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: MCCARTHY

DATE/TIME OF TEST: 09/14/2019 02:43:55

Diagnostics Check	OK	02:36
Air Blank	0.000	02:36
Control Test	0.079	02:36
Air Blank	0.000	02:37
Subject Sample # 1	0.187	02:39
Breath Volume	1.546	02:39
Air Blank	0.000	02:39
Air Blank	0.000	02:41
Subject Sample # 2	0.190	02:42
Breath Volume	3.042	02:42
Air Blank	0.000	02:42
Control Test	0.078	02:43
Air Blank	0.000	02:43
Diagnostics Check	OK	02:43

Data Download Date/Time: 09/27/2019 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001155

INSTRUMENT REGISTERED TO: EGLIN AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAKUBOWSKI,SCOTT,T
BREATH TEST OPERATOR AGENCY: Eglin AFB PD
ARRESTING OFFICER: NORMAN,WILLIAM
ARRESTING OFFICER AGENCY: Eglin AFB PD
LAST AGENCY INSPECTION: 08/12/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: ZERO TOL
ARREST TIME: 03:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:59

SUBJECT'S LAST NAME: DANIELS

DATE/TIME OF TEST: 09/07/2019 05:38:17

Diagnostics Check	OK	05:28
Air Blank	0.000	05:28
Control Test	0.079	05:29
Air Blank	0.000	05:29
Subject Sample # 1	0.107	05:32
Breath Volume	1.597	05:32
Air Blank	0.000	05:33
Air Blank	0.000	05:35
Subject Sample # 2	0.097	05:36
Breath Volume	1.816	05:36
Air Blank	0.000	05:37
Control Test	0.077	05:37
Air Blank	0.000	05:38
Diagnostics Check	OK	05:38

Data Download Date/Time: 09/17/2019 12:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001155

INSTRUMENT REGISTERED TO: EGLIN AFB

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: JAKUBOWSKI,SCOTT,T

BREATH TEST OPERATOR AGENCY: Eglin AFB PD

ARRESTING OFFICER: WELLER,ZACHARY

ARRESTING OFFICER AGENCY: Eglin AFB PD

LAST AGENCY INSPECTION: 08/12/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 05218080A4

DRY GAS EXPIRY DATE: 05/05/2020

VIOLATION CODE: ZERO TOL

ARREST TIME: 03:57

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:59

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 09/07/2019 05:58:07

Diagnostics Check	OK	05:51
Air Blank	0.000	05:51
Control Test	0.078	05:52
Air Blank	0.000	05:52
Subject Sample # 1	0.051	05:53
Breath Volume	1.121	05:53
Air Blank	0.000	05:53
Air Blank	0.000	05:55
Subject Sample # 2	0.052	05:56
Breath Volume	2.890	05:56
Air Blank	0.000	05:57
Control Test	0.078	05:57
Air Blank	0.000	05:58
Diagnostics Check	OK	05:58

Data Download Date/Time: 09/17/2019 12:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001155

INSTRUMENT REGISTERED TO: EGLIN AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASEY,JOSHUA,P
BREATH TEST OPERATOR AGENCY: Eglin AFB PD
ARRESTING OFFICER: CASEY,JOSHUA
ARRESTING OFFICER AGENCY: Eglin AFB PD
LAST AGENCY INSPECTION: 09/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 09/13/2019 05:31:16

Diagnostics Check	OK	05:20	
Air Blank	0.000	05:21	
Control Test	0.079	05:21	
Air Blank	0.000	05:22	
Subject Sample	# 1 0.163	05:23	
Breath Volume	1.546	05:23	
Air Blank	0.000	05:23	
Air Blank	0.000	05:25	
Subject Sample	# 2 0.123	05:26	
Breath Volume	2.394	05:26	
Air Blank	0.000	05:27	
Air Blank	0.000	05:28	
Subject Sample	# 3 0.128	05:29	
Breath Volume	2.386	05:29	
Air Blank	0.000	05:30	
Control Test	0.072	05:30	Control Outside Tolerance
Air Blank	0.000	05:31	

Data Download Date/Time: 09/17/2019 12:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001155

INSTRUMENT REGISTERED TO: EGLIN AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAKUBOWSKI,SCOTT,T
BREATH TEST OPERATOR AGENCY: Eglin AFB PD
ARRESTING OFFICER: JAKUBOWSKI,SCOTT
ARRESTING OFFICER AGENCY: Eglin AFB PD
LAST AGENCY INSPECTION: 09/17/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: ZERO TOL
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 09/22/2019 02:28:17

Diagnostics Check	OK	02:22	
Air Blank	0.000	02:23	
Control Test	0.078	02:23	
Air Blank	0.000	02:24	
Subject Sample	# 1 0.181	02:24	
Breath Volume	2.070	02:24	
Air Blank	0.000	02:25	
Air Blank	0.000	02:27	
Subject Sample	# 2 0.183	02:27	
Breath Volume	1.210	02:27	
Air Blank		02:28	RFI Detect
Air Blank		02:28	RFI Detect

Data Download Date/Time: 10/09/2019 13:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,RICHARD,L
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: FOWLER,JESSICA
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 08/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 17:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:02

SUBJECT'S LAST NAME: CAGGIANO

DATE/TIME OF TEST: 09/12/2019 18:29:38

Diagnostics Check	OK	18:26	
Air Blank	0.000	18:27	
Control Test	0.080	18:27	
Air Blank	0.000	18:28	
Subject Sample	# 1	18:28	Subject Test Refused
Breath Volume	0.000	18:28	
Air Blank	0.000	18:28	
Control Test	0.081	18:29	
Air Blank	0.000	18:29	
Diagnostics Check	OK	18:29	

Data Download Date/Time: 09/29/2019 11:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUBIN,RICHARD,D
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: RUBIN,RICHARD
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 08/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: HEIGHLEY

DATE/TIME OF TEST: 09/15/2019 04:51:57

Diagnostics Check	OK	04:40	
Air Blank	0.000	04:40	
Control Test	0.081	04:41	
Air Blank	0.000	04:41	
Subject Sample	# 1 0.169	04:44	Volume Not Met
Breath Volume	0.953	04:44	
Air Blank	0.000	04:45	
Air Blank	0.000	04:47	
Subject Sample	# 2 0.171	04:50	Volume Not Met
Breath Volume	0.355	04:50	
Air Blank	0.000	04:51	
Control Test	0.081	04:51	
Air Blank	0.000	04:51	
Diagnostics Check	OK	04:51	

Data Download Date/Time: 09/29/2019 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,RICHARD,L
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: FOWLER,JESSICA
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 08/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: BROWNELL

DATE/TIME OF TEST: 09/22/2019 03:52:29

Diagnostics Check	OK	03:45
Air Blank	0.000	03:46
Control Test	0.081	03:46
Air Blank	0.000	03:46
Subject Sample # 1	0.153	03:47
Breath Volume	4.148	03:47
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 2	0.144	03:51
Breath Volume	3.031	03:51
Air Blank	0.000	03:51
Control Test	0.081	03:51
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time: 09/29/2019 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,RICHARD,L
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: SCOTT,KARI
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 08/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 16:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:13

SUBJECT'S LAST NAME: PHELAN

DATE/TIME OF TEST: 09/25/2019 17:49:20

Diagnostics Check	OK	17:37	
Air Blank	0.000	17:38	
Control Test	0.081	17:38	
Air Blank	0.000	17:38	
Subject Sample	# 1 0.164	17:42	Volume Not Met
Breath Volume	0.296	17:42	
Air Blank	0.000	17:42	
Air Blank	0.000	17:44	
Subject Sample	# 2 0.153	17:47	Volume Not Met
Breath Volume	0.769	17:47	
Air Blank	0.000	17:48	
Control Test	0.081	17:48	
Air Blank	0.000	17:49	
Diagnostics Check	OK	17:49	

Data Download Date/Time: 09/29/2019 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,RICHARD,L
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: SCOTT,KARI
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 08/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 16:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:13

SUBJECT'S LAST NAME: PHELAN

DATE/TIME OF TEST: 09/25/2019 18:06:46

Diagnostics Check	OK	17:52	
Air Blank	0.000	17:52	
Control Test	0.079	17:52	
Air Blank	0.000	17:53	
Subject Sample	# 1 0.167	17:56	Volume Not Met
Breath Volume	0.460	17:56	
Air Blank	0.000	17:57	
Air Blank	0.000	17:58	
Subject Sample	# 2 0.170	18:01	
Breath Volume	1.433	18:01	
Air Blank	0.000	18:02	
Air Blank	0.000	18:04	
Subject Sample	# 3 0.164	18:05	
Breath Volume	1.523	18:05	
Air Blank	0.000	18:05	
Control Test	0.078	18:06	
Air Blank	0.000	18:06	
Diagnostics Check	OK	18:06	

Data Download Date/Time: 09/29/2019 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001159

INSTRUMENT REGISTERED TO: ORANGE CITY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUBAKER,KELLY,J
BREATH TEST OPERATOR AGENCY: Orange City PD
ARRESTING OFFICER: KING,BRANDON
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 09/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: CARROLL

DATE/TIME OF TEST: 09/07/2019 03:18:27

Diagnostics Check	OK	03:10
Air Blank	0.000	03:10
Control Test	0.082	03:10
Air Blank	0.000	03:11
Subject Sample # 1	0.211	03:13
Breath Volume	1.703	03:13
Air Blank	0.000	03:14
Air Blank	0.000	03:16
Subject Sample # 2	0.229	03:16
Breath Volume	1.898	03:16
Air Blank	0.000	03:17
Control Test	0.081	03:17
Air Blank	0.000	03:18
Diagnostics Check	OK	03:18

Data Download Date/Time: 11/14/2019 12:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001159

INSTRUMENT REGISTERED TO: ORANGE CITY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRALSON,SAMUEL,E
BREATH TEST OPERATOR AGENCY: Orange City PD
ARRESTING OFFICER: CROFUTT,BRITTANY
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 09/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: ROEBUCK

DATE/TIME OF TEST: 09/28/2019 03:36:11

Diagnostics Check	OK	03:29
Air Blank	0.000	03:30
Control Test	0.081	03:30
Air Blank	0.000	03:31
Subject Sample # 1	0.109	03:31
Breath Volume	3.570	03:31
Air Blank	0.000	03:32
Air Blank	0.000	03:34
Subject Sample # 2	0.110	03:34
Breath Volume	3.925	03:34
Air Blank	0.000	03:35
Control Test	0.081	03:35
Air Blank	0.000	03:36
Diagnostics Check	OK	03:36

Data Download Date/Time: 11/14/2019 12:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001159

INSTRUMENT REGISTERED TO: ORANGE CITY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRALSON,SAMUEL,E
BREATH TEST OPERATOR AGENCY: Orange City PD
ARRESTING OFFICER: JAMES,ROYCE
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 09/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: GALVAN

DATE/TIME OF TEST: 09/28/2019 04:08:56

Diagnostics Check	OK	04:01
Air Blank	0.000	04:02
Control Test	0.081	04:02
Air Blank	0.000	04:03
Subject Sample # 1	0.192	04:04
Breath Volume	1.500	04:04
Air Blank	0.000	04:04
Air Blank	0.000	04:06
Subject Sample # 2	0.203	04:07
Breath Volume	1.644	04:07
Air Blank	0.000	04:08
Control Test	0.080	04:08
Air Blank	0.000	04:08
Diagnostics Check	OK	04:08

Data Download Date/Time: 11/14/2019 12:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRUSE, KRISTOPHER, D
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: PARISH, CLINTON
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 08/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: RIVERA RIVERA

DATE/TIME OF TEST: 09/07/2019 01:51:53

Diagnostics Check	OK	01:39	
Air Blank	0.000	01:40	
Control Test	0.080	01:40	
Air Blank	0.000	01:41	
Subject Sample	# 1	01:44	No Sample Provided
Breath Volume	0.000	01:44	
Air Blank	0.000	01:44	
Air Blank	0.000	01:46	
Subject Sample	# 2	0.214	01:47
Breath Volume	2.617	01:47	
Air Blank	0.000	01:47	
Air Blank	0.000	01:49	
Subject Sample	# 3	0.216	01:50
Breath Volume	1.980	01:50	
Air Blank	0.000	01:50	
Control Test	0.080	01:51	
Air Blank	0.000	01:51	
Diagnostics Check	OK	01:51	

Data Download Date/Time: 09/26/2019 15:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRUSE, KRISTOPHER, D
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: COONS, SEAN
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 08/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: KING-HOLDER

DATE/TIME OF TEST: 09/23/2019 04:37:02

Diagnostics Check	OK	04:31	
Air Blank	0.000	04:31	
Control Test	0.080	04:32	
Air Blank	0.000	04:32	
Subject Sample	# 1	04:35	RFI Detect
Breath Volume	1.222	04:35	
Air Blank	0.000	04:36	
Control Test	0.081	04:36	
Air Blank	0.000	04:36	
Diagnostics Check	OK	04:37	

Data Download Date/Time: 09/26/2019 15:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KRUSE, KRISTOPHER, D

BREATH TEST OPERATOR AGENCY: Clermont PD

ARRESTING OFFICER: COONS, SEAN

ARRESTING OFFICER AGENCY: Clermont PD

LAST AGENCY INSPECTION: 08/01/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 02618080A2

DRY GAS EXPIRY DATE: 03/05/2020

VIOLATION CODE: DUI

ARREST TIME: 04:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: KING-HOLDER

DATE/TIME OF TEST: 09/23/2019 04:45:40

Diagnostics Check	OK	04:39
Air Blank	0.000	04:39
Control Test	0.079	04:39
Air Blank	0.000	04:40
Subject Sample # 1	0.149	04:41
Breath Volume	2.078	04:41
Air Blank	0.000	04:41
Air Blank	0.000	04:43
Subject Sample # 2	0.150	04:44
Breath Volume	1.589	04:44
Air Blank	0.000	04:44
Control Test	0.080	04:45
Air Blank	0.000	04:45
Diagnostics Check	OK	04:45

Data Download Date/Time: 09/26/2019 15:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HATT,CODY,L
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: HATT,CODY
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 09/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 13:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:20

SUBJECT'S LAST NAME: BATTLE

DATE/TIME OF TEST: 09/23/2019 16:52:12

Diagnostics Check	OK	16:45
Air Blank	0.000	16:45
Control Test	0.080	16:45
Air Blank	0.000	16:46
Subject Sample # 1	0.050	16:47
Breath Volume	2.941	16:47
Air Blank	0.000	16:48
Air Blank	0.000	16:50
Subject Sample # 2	0.042	16:50
Breath Volume	1.949	16:50
Air Blank	0.000	16:51
Control Test	0.079	16:51
Air Blank	0.000	16:52
Diagnostics Check	OK	16:52

Data Download Date/Time: 10/04/2019 17:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001190

INSTRUMENT REGISTERED TO: BOYNTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTRO,DENNIS,
BREATH TEST OPERATOR AGENCY: Boynton Beach PD
ARRESTING OFFICER: CASTRO,DENNIS
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 09/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: PAES DE CARVALHO BAE

DATE/TIME OF TEST: 09/06/2019 03:41:14

Diagnostics Check	OK	03:35
Air Blank	0.000	03:35
Control Test	0.077	03:35
Air Blank	0.000	03:36
Subject Sample # 1	0.207	03:36
Breath Volume	3.417	03:36
Air Blank	0.000	03:37
Air Blank	0.000	03:39
Subject Sample # 2	0.191	03:39
Breath Volume	1.347	03:39
Air Blank	0.000	03:40
Control Test	0.076	03:40
Air Blank	0.000	03:41
Diagnostics Check	OK	03:41

Data Download Date/Time: 10/09/2019 17:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001190

INSTRUMENT REGISTERED TO: BOYNTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTRO,DENNIS,
BREATH TEST OPERATOR AGENCY: Boynton Beach PD
ARRESTING OFFICER: CASTRO,DENNIS
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 09/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: MCLEOD

DATE/TIME OF TEST: 09/20/2019 21:52:35

Diagnostics Check	OK	21:49	
Air Blank	0.000	21:50	
Control Test	0.077	21:50	
Air Blank	0.000	21:51	
Subject Sample	# 1	21:51	Subject Test Refused
Breath Volume	0.000	21:51	
Air Blank	0.000	21:51	
Control Test	0.077	21:51	
Air Blank	0.000	21:52	
Diagnostics Check	OK	21:52	

Data Download Date/Time: 10/09/2019 17:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARNOLD,J,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: ARNOLD,J
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 09/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:11

SUBJECT'S LAST NAME: TRICE

DATE/TIME OF TEST: 09/04/2019 22:43:37

Diagnostics Check	OK	22:37
Air Blank	0.000	22:37
Control Test	0.079	22:38
Air Blank	0.000	22:38
Subject Sample # 1	0.286	22:39
Breath Volume	1.992	22:39
Air Blank	0.000	22:39
Air Blank	0.000	22:41
Subject Sample # 2	0.280	22:42
Breath Volume	1.839	22:42
Air Blank	0.000	22:42
Control Test	0.078	22:43
Air Blank	0.000	22:43
Diagnostics Check	OK	22:43

Data Download Date/Time: 10/03/2019 09:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORNELIUS,TIM,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: CORNELIUS,TIM
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 09/03/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:28

SUBJECT'S LAST NAME: BALLOU

DATE/TIME OF TEST: 09/06/2019 21:58:59

Diagnostics Check	OK	21:51
Air Blank	0.000	21:52
Control Test	0.078	21:52
Air Blank	0.000	21:53
Subject Sample # 1	0.103	21:54
Breath Volume	2.109	21:54
Air Blank	0.000	21:54
Air Blank	0.000	21:56
Subject Sample # 2	0.109	21:57
Breath Volume	1.433	21:57
Air Blank	0.000	21:58
Control Test	0.078	21:58
Air Blank	0.000	21:58
Diagnostics Check	OK	21:58

Data Download Date/Time: 10/03/2019 09:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,L
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: KENNEDY,DANIEL
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 09/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: PULVER

DATE/TIME OF TEST: 09/07/2019 01:06:23

Diagnostics Check	OK	00:59
Air Blank	0.000	00:59
Control Test	0.078	01:00
Air Blank	0.000	01:00
Subject Sample # 1	0.171	01:01
Breath Volume	1.812	01:01
Air Blank	0.000	01:01
Air Blank	0.000	01:03
Subject Sample # 2	0.162	01:04
Breath Volume	1.394	01:04
Air Blank	0.000	01:05
Control Test	0.077	01:05
Air Blank	0.000	01:06
Diagnostics Check	OK	01:06

Data Download Date/Time: 10/03/2019 09:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,E,W
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: BRAWNER,M
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 09/03/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: SHORT

DATE/TIME OF TEST: 09/07/2019 20:47:51

Diagnostics Check	OK	20:41
Air Blank	0.000	20:42
Control Test	0.078	20:42
Air Blank	0.000	20:43
Subject Sample # 1	0.290	20:43
Breath Volume	2.273	20:43
Air Blank	0.000	20:44
Air Blank	0.000	20:45
Subject Sample # 2	0.289	20:46
Breath Volume	2.457	20:46
Air Blank	0.000	20:46
Control Test	0.078	20:47
Air Blank	0.000	20:47
Diagnostics Check	OK	20:47

Data Download Date/Time: 10/03/2019 09:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARP,RYAN,M
BREATH TEST OPERATOR AGENCY: NAPLES PD
ARRESTING OFFICER: HARP,RYAN
ARRESTING OFFICER AGENCY: NAPLES PD
LAST AGENCY INSPECTION: 09/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: FRASER

DATE/TIME OF TEST: 09/11/2019 02:25:48

Diagnostics Check	OK	02:19
Air Blank	0.000	02:19
Control Test	0.078	02:20
Air Blank	0.000	02:20
Subject Sample # 1	0.136	02:21
Breath Volume	1.996	02:21
Air Blank	0.000	02:21
Air Blank	0.000	02:23
Subject Sample # 2	0.126	02:24
Breath Volume	1.785	02:24
Air Blank	0.000	02:24
Control Test	0.078	02:25
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

Data Download Date/Time: 10/03/2019 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,L
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: KENNEDY,DANIEL
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 09/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:58

SUBJECT'S LAST NAME: SPINELLA

DATE/TIME OF TEST: 09/11/2019 18:30:34

Diagnostics Check	OK	18:24
Air Blank	0.000	18:24
Control Test	0.078	18:25
Air Blank	0.000	18:25
Subject Sample # 1	0.000	18:26
Breath Volume	1.355	18:26
Air Blank	0.000	18:26
Air Blank	0.000	18:28
Subject Sample # 2	0.000	18:29
Breath Volume	2.445	18:29
Air Blank	0.000	18:29
Control Test	0.078	18:29
Air Blank	0.000	18:30
Diagnostics Check	OK	18:30

Data Download Date/Time: 10/03/2019 09:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DECKER,DAVID,P
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: TORRES,B
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 09/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 11:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:45

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 09/12/2019 13:22:43

Diagnostics Check	OK	13:14
Air Blank	0.000	13:15
Control Test	0.079	13:15
Air Blank	0.000	13:16
Subject Sample # 1	0.000	13:18
Breath Volume	3.062	13:18
Air Blank	0.000	13:18
Air Blank	0.000	13:20
Subject Sample # 2	0.000	13:21
Breath Volume	3.167	13:21
Air Blank	0.000	13:21
Control Test	0.078	13:22
Air Blank	0.000	13:22
Diagnostics Check	OK	13:22

Data Download Date/Time: 10/03/2019 09:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKS,D,M
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: MILLER,K
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 09/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: OBRIEN

DATE/TIME OF TEST: 09/12/2019 23:08:04

Diagnostics Check	OK	23:01
Air Blank	0.000	23:02
Control Test	0.078	23:02
Air Blank	0.000	23:03
Subject Sample # 1	0.139	23:03
Breath Volume	2.109	23:03
Air Blank	0.000	23:04
Air Blank	0.000	23:06
Subject Sample # 2	0.137	23:06
Breath Volume	1.238	23:06
Air Blank	0.000	23:07
Control Test	0.078	23:07
Air Blank	0.000	23:07
Diagnostics Check	OK	23:08

Data Download Date/Time: 10/03/2019 09:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOLIEK,P,A
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: GARCIA,R
ARRESTING OFFICER AGENCY: FHP TROOP F
LAST AGENCY INSPECTION: 09/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:31

SUBJECT'S LAST NAME: VALENTINE

DATE/TIME OF TEST: 09/13/2019 17:00:54

Diagnostics Check	OK	16:54
Air Blank	0.000	16:54
Control Test	0.079	16:55
Air Blank	0.000	16:55
Subject Sample # 1	0.205	16:56
Breath Volume	2.917	16:56
Air Blank	0.000	16:57
Air Blank	0.000	16:58
Subject Sample # 2	0.209	16:59
Breath Volume	1.839	16:59
Air Blank	0.000	17:00
Control Test	0.078	17:00
Air Blank	0.000	17:00
Diagnostics Check	OK	17:00

Data Download Date/Time: 10/03/2019 09:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVAS,W,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: APONTE,C
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 09/03/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:42

SUBJECT'S LAST NAME: CARRERA

DATE/TIME OF TEST: 09/13/2019 20:10:03

Diagnostics Check	OK	20:02	
Air Blank	0.000	20:02	
Control Test	0.078	20:03	
Air Blank	0.000	20:03	
Subject Sample	# 1 0.277	20:03	
Breath Volume	1.687	20:03	
Air Blank	0.000	20:04	
Air Blank	0.000	20:06	
Subject Sample	# 2	20:08	Subject Test Refused
Breath Volume	0.000	20:08	
Air Blank	0.000	20:09	
Control Test	0.078	20:09	
Air Blank	0.000	20:09	
Diagnostics Check	OK	20:10	

Data Download Date/Time: 10/03/2019 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUGO,JONATHAN,B
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: STEVENSON,C
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 09/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: ORRISON

DATE/TIME OF TEST: 09/14/2019 02:03:35

Diagnostics Check	OK	01:56	
Air Blank	0.000	01:56	
Control Test	0.079	01:56	
Air Blank	0.000	01:57	
Subject Sample # 1	0.184	01:58	
Breath Volume	1.292	01:58	
Air Blank	0.000	01:59	
Air Blank	0.000	02:00	
Subject Sample # 2	0.172	02:02	
Breath Volume	1.285	02:02	
Air Blank	0.000	02:02	
Control Test	0.066	02:03	Control Outside Tolerance
Air Blank	0.000	02:03	

Data Download Date/Time: 10/03/2019 09:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUGO,JONATHAN,B
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: STEVENSON,C
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 09/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: ORRISON

DATE/TIME OF TEST: 09/14/2019 02:13:16

Diagnostics Check	OK	02:06
Air Blank	0.000	02:07
Control Test	0.078	02:07
Air Blank	0.000	02:07
Subject Sample # 1	0.182	02:08
Breath Volume	1.304	02:08
Air Blank	0.000	02:09
Air Blank	0.000	02:11
Subject Sample # 2	0.165	02:11
Breath Volume	1.554	02:11
Air Blank	0.000	02:12
Control Test	0.078	02:12
Air Blank	0.000	02:13
Diagnostics Check	OK	02:13

Data Download Date/Time: 10/03/2019 09:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOPS,C,M
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: ROWLEY,C
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 09/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: LAVIGNE

DATE/TIME OF TEST: 09/14/2019 04:28:07

Diagnostics Check	OK	04:22
Air Blank	0.000	04:22
Control Test	0.078	04:22
Air Blank	0.000	04:23
Subject Sample # 1	0.103	04:23
Breath Volume	3.011	04:23
Air Blank	0.000	04:24
Air Blank	0.000	04:26
Subject Sample # 2	0.103	04:26
Breath Volume	2.968	04:26
Air Blank	0.000	04:27
Control Test	0.078	04:27
Air Blank	0.000	04:28
Diagnostics Check	OK	04:28

Data Download Date/Time: 10/03/2019 09:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,S,C
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: GROSSKLAS,D
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 09/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:34

SUBJECT'S LAST NAME: CARDOSO

DATE/TIME OF TEST: 09/15/2019 22:03:30

Diagnostics Check	OK	21:57
Air Blank	0.000	21:57
Control Test	0.079	21:57
Air Blank	0.000	21:58
Subject Sample # 1	0.146	21:58
Breath Volume	1.593	21:58
Air Blank	0.000	21:59
Air Blank	0.000	22:01
Subject Sample # 2	0.141	22:02
Breath Volume	1.601	22:02
Air Blank	0.000	22:02
Control Test	0.078	22:02
Air Blank	0.000	22:03
Diagnostics Check	OK	22:03

Data Download Date/Time: 10/03/2019 09:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORNELIUS,TIM,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: SMITH,LORENZO
ARRESTING OFFICER AGENCY: MARCO ISLAND PD
LAST AGENCY INSPECTION: 09/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: VAIL

DATE/TIME OF TEST: 09/17/2019 01:27:39

Diagnostics Check	OK	01:21
Air Blank	0.000	01:21
Control Test	0.078	01:22
Air Blank	0.000	01:22
Subject Sample # 1	0.000	01:23
Breath Volume	2.261	01:23
Air Blank	0.000	01:23
Air Blank	0.000	01:25
Subject Sample # 2	0.000	01:26
Breath Volume	2.519	01:26
Air Blank	0.000	01:26
Control Test	0.078	01:27
Air Blank	0.000	01:27
Diagnostics Check	OK	01:27

Data Download Date/Time: 10/03/2019 09:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHLINE,C,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: VAUGHAN,B
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 09/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: DETTLOFF

DATE/TIME OF TEST: 09/17/2019 21:38:54

Diagnostics Check	OK	21:31
Air Blank	0.000	21:31
Control Test	0.079	21:31
Air Blank	0.000	21:32
Subject Sample # 1	0.053	21:33
Breath Volume	1.191	21:33
Air Blank	0.000	21:33
Air Blank	0.000	21:35
Subject Sample # 2	0.051	21:37
Breath Volume	1.359	21:37
Air Blank	0.000	21:37
Control Test	0.078	21:38
Air Blank	0.000	21:38
Diagnostics Check	OK	21:38

Data Download Date/Time: 10/03/2019 09:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEORGE,SHANNON,N
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: GINTHER,D
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 09/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: PARRA

DATE/TIME OF TEST: 09/19/2019 04:13:48

Diagnostics Check	OK	04:05
Air Blank	0.000	04:06
Control Test	0.078	04:06
Air Blank	0.000	04:07
Subject Sample # 1	0.143	04:09
Breath Volume	1.531	04:09
Air Blank	0.000	04:09
Air Blank	0.000	04:11
Subject Sample # 2	0.150	04:12
Breath Volume	1.515	04:12
Air Blank	0.000	04:12
Control Test	0.078	04:13
Air Blank	0.000	04:13
Diagnostics Check	OK	04:13

Data Download Date/Time: 10/03/2019 09:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,L
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: KENNEDY,DANIEL
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 09/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:16

SUBJECT'S LAST NAME: IXCOY CHAVEZ

DATE/TIME OF TEST: 09/21/2019 21:46:52

Diagnostics Check	OK	21:39
Air Blank	0.000	21:39
Control Test	0.079	21:40
Air Blank	0.000	21:40
Subject Sample # 1	0.148	21:42
Breath Volume	2.640	21:42
Air Blank	0.000	21:42
Air Blank	0.000	21:44
Subject Sample # 2	0.134	21:45
Breath Volume	1.246	21:45
Air Blank	0.000	21:45
Control Test	0.078	21:46
Air Blank	0.000	21:46
Diagnostics Check	OK	21:46

Data Download Date/Time: 10/03/2019 09:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,L
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: KENNEDY,DANIEL
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 09/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:26

SUBJECT'S LAST NAME: CHESTER

DATE/TIME OF TEST: 09/22/2019 17:56:02

Diagnostics Check	OK	17:49
Air Blank	0.000	17:50
Control Test	0.079	17:50
Air Blank	0.000	17:50
Subject Sample # 1	0.031	17:51
Breath Volume	2.027	17:51
Air Blank	0.000	17:52
Air Blank	0.000	17:54
Subject Sample # 2	0.031	17:54
Breath Volume	2.199	17:54
Air Blank	0.000	17:55
Control Test	0.079	17:55
Air Blank	0.000	17:55
Diagnostics Check	OK	17:56

Data Download Date/Time: 10/03/2019 09:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANNORAT,ADELY,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: BRAWNER,M
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 09/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: CHECA

DATE/TIME OF TEST: 09/23/2019 00:44:00

Diagnostics Check	OK	00:37
Air Blank	0.000	00:37
Control Test	0.079	00:37
Air Blank	0.000	00:38
Subject Sample # 1	0.179	00:39
Breath Volume	1.355	00:39
Air Blank	0.000	00:39
Air Blank	0.000	00:41
Subject Sample # 2	0.197	00:42
Breath Volume	1.449	00:42
Air Blank	0.000	00:43
Control Test	0.079	00:43
Air Blank	0.000	00:43
Diagnostics Check	OK	00:44

Data Download Date/Time: 10/03/2019 09:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EMANS,ALEX,R
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: WOODS,H
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 09/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: GOSSARD

DATE/TIME OF TEST: 09/23/2019 02:10:58

Diagnostics Check	OK	02:04
Air Blank	0.000	02:05
Control Test	0.079	02:05
Air Blank	0.000	02:06
Subject Sample # 1	0.174	02:06
Breath Volume	1.480	02:06
Air Blank	0.000	02:07
Air Blank	0.000	02:08
Subject Sample # 2	0.182	02:09
Breath Volume	1.917	02:09
Air Blank	0.000	02:10
Control Test	0.078	02:10
Air Blank	0.000	02:10
Diagnostics Check	OK	02:10

Data Download Date/Time: 10/03/2019 09:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORNELIUS,TIM,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: CORNELIUS,TIM
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 09/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: HEATH

DATE/TIME OF TEST: 09/25/2019 23:00:03

Diagnostics Check	OK	22:53
Air Blank	0.000	22:53
Control Test	0.079	22:54
Air Blank	0.000	22:54
Subject Sample # 1	0.209	22:55
Breath Volume	3.972	22:55
Air Blank	0.000	22:55
Air Blank	0.000	22:57
Subject Sample # 2	0.210	22:58
Breath Volume	4.281	22:58
Air Blank	0.000	22:59
Control Test	0.078	22:59
Air Blank	0.000	22:59
Diagnostics Check	OK	23:00

Data Download Date/Time: 10/03/2019 09:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIERMAN,B,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: VAUGHAN,B
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 09/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: BERTOLOTTI

DATE/TIME OF TEST: 09/26/2019 04:29:29

Diagnostics Check	OK	04:23
Air Blank	0.000	04:23
Control Test	0.078	04:24
Air Blank	0.000	04:24
Subject Sample # 1	0.215	04:24
Breath Volume	1.113	04:24
Air Blank	0.000	04:25
Air Blank	0.000	04:27
Subject Sample # 2	0.217	04:27
Breath Volume	1.523	04:27
Air Blank	0.000	04:28
Control Test	0.078	04:28
Air Blank	0.000	04:29
Diagnostics Check	OK	04:29

Data Download Date/Time: 10/03/2019 09:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONGORA,ANTONIO,A
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: BURD,S
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 09/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 09:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:25

SUBJECT'S LAST NAME: COPELAND

DATE/TIME OF TEST: 09/26/2019 13:54:06

Diagnostics Check	OK	13:47
Air Blank	0.000	13:48
Control Test	0.078	13:48
Air Blank	0.000	13:49
Subject Sample # 1	0.000	13:49
Breath Volume	1.214	13:49
Air Blank	0.000	13:50
Air Blank	0.000	13:52
Subject Sample # 2	0.000	13:52
Breath Volume	1.527	13:52
Air Blank	0.000	13:53
Control Test	0.078	13:53
Air Blank	0.000	13:54
Diagnostics Check	OK	13:54

Data Download Date/Time: 10/03/2019 09:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOODMAN,R.,R
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: BAXTER,J.
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 09/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 13:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:11

SUBJECT'S LAST NAME: KILLILEA

DATE/TIME OF TEST: 09/26/2019 15:38:40

Diagnostics Check	OK	15:35	
Air Blank	0.000	15:35	
Control Test	0.078	15:36	
Air Blank	0.000	15:36	
Subject Sample	# 1	15:37	Interferent Detect
Breath Volume	3.039	15:37	
Air Blank	0.000	15:37	
Control Test	0.078	15:38	
Air Blank	0.000	15:38	
Diagnostics Check	OK	15:38	

Data Download Date/Time: 10/03/2019 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOODMAN,ROBERT,R
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: BAXTER,J.
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 09/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 13:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:11

SUBJECT'S LAST NAME: KILLILEA

DATE/TIME OF TEST: 09/26/2019 15:44:00

Diagnostics Check	OK	15:40	
Air Blank	0.000	15:40	
Control Test	0.078	15:41	
Air Blank	0.000	15:41	
Subject Sample	# 1	15:42	Interferent Detect
Breath Volume	3.574	15:42	
Air Blank	0.000	15:43	
Control Test	0.079	15:43	
Air Blank	0.000	15:43	
Diagnostics Check	OK	15:44	

Data Download Date/Time: 10/03/2019 09:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,THOMAS,L
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: GINTHER,D
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 09/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 09/28/2019 02:07:19

Diagnostics Check	OK	02:00
Air Blank	0.000	02:01
Control Test	0.078	02:01
Air Blank	0.000	02:02
Subject Sample # 1	0.128	02:02
Breath Volume	3.386	02:02
Air Blank	0.000	02:03
Air Blank	0.000	02:05
Subject Sample # 2	0.118	02:05
Breath Volume	3.546	02:05
Air Blank	0.000	02:06
Control Test	0.078	02:06
Air Blank	0.000	02:07
Diagnostics Check	OK	02:07

Data Download Date/Time: 10/03/2019 09:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELANEY,PATRICK,T
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: FLEURANT,M
ARRESTING OFFICER AGENCY: FHP TROOP F
LAST AGENCY INSPECTION: 09/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 13:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:33

SUBJECT'S LAST NAME: LUSSIER

DATE/TIME OF TEST: 09/28/2019 15:06:07

Diagnostics Check	OK	14:58	
Air Blank	0.000	14:58	
Control Test	0.079	14:58	
Air Blank	0.000	14:59	
Subject Sample	# 1 0.011	15:00	
Breath Volume	1.609	15:00	
Air Blank	0.000	15:01	
Air Blank	0.000	15:03	
Subject Sample	# 2 0.009	15:04	
Breath Volume	1.539	15:04	
Air Blank	0.000	15:04	
Control Test	0.077	15:05	
Air Blank		15:05	RFI Detect
Air Blank	0.000	15:06	

Data Download Date/Time: 10/03/2019 09:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODI, DOMENIC,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: ZERPA, E
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 09/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: SHELTON

DATE/TIME OF TEST: 09/29/2019 22:59:34

Diagnostics Check	OK	22:53
Air Blank	0.000	22:53
Control Test	0.079	22:54
Air Blank	0.000	22:54
Subject Sample # 1	0.141	22:55
Breath Volume	2.644	22:55
Air Blank	0.000	22:55
Air Blank	0.000	22:57
Subject Sample # 2	0.139	22:58
Breath Volume	2.664	22:58
Air Blank	0.000	22:58
Control Test	0.079	22:58
Air Blank	0.000	22:59
Diagnostics Check	OK	22:59

Data Download Date/Time: 10/03/2019 09:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABRAHAM,ERIC,D
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: MCNEIL,RAYMOND
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 08/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: WATTS

DATE/TIME OF TEST: 09/02/2019 03:27:35

Diagnostics Check	OK	03:20
Air Blank	0.000	03:21
Control Test	0.080	03:21
Air Blank	0.000	03:22
Subject Sample # 1	0.096	03:23
Breath Volume	3.777	03:23
Air Blank	0.000	03:23
Air Blank	0.000	03:25
Subject Sample # 2	0.093	03:26
Breath Volume	3.070	03:26
Air Blank	0.000	03:26
Control Test	0.080	03:27
Air Blank	0.000	03:27
Diagnostics Check	OK	03:27

Data Download Date/Time: 09/04/2019 04:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABRAHAM,ERIC,D
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: ABRAHAM,ERIC
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 08/29/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 04:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: BRANSCOMB

DATE/TIME OF TEST: 09/02/2019 05:22:10

Diagnostics Check	OK	05:11
Air Blank	0.000	05:12
Control Test	0.079	05:12
Air Blank	0.000	05:13
Subject Sample # 1	0.210	05:13
Breath Volume	1.582	05:13
Air Blank	0.000	05:14
Air Blank	0.000	05:16
Subject Sample # 2	0.246	05:17
Breath Volume	1.679	05:17
Air Blank	0.000	05:17
Air Blank	0.000	05:19
Subject Sample # 3	0.234	05:20
Breath Volume	1.792	05:20
Air Blank	0.000	05:21
Control Test	0.080	05:21
Air Blank	0.000	05:22
Diagnostics Check	OK	05:22

Data Download Date/Time: 09/04/2019 04:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,JONATHAN,R
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: JOHNSON,MICHAEL
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 08/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 09/06/2019 02:31:39

Diagnostics Check	OK	02:28	
Air Blank	0.000	02:29	
Control Test	0.079	02:29	
Air Blank	0.000	02:30	
Subject Sample	# 1	02:30	Subject Test Refused
Breath Volume	0.000	02:30	
Air Blank	0.000	02:30	
Control Test	0.078	02:31	
Air Blank	0.000	02:31	
Diagnostics Check	OK	02:31	

Data Download Date/Time: 10/02/2019 11:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,JONATHAN,R
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: ALLYN,PRESTON
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 08/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: SUAREZ ALMEIDA

DATE/TIME OF TEST: 09/06/2019 02:39:31

Diagnostics Check	OK	02:33
Air Blank	0.000	02:33
Control Test	0.079	02:34
Air Blank	0.000	02:34
Subject Sample # 1	0.210	02:35
Breath Volume	1.589	02:35
Air Blank	0.000	02:35
Air Blank	0.000	02:37
Subject Sample # 2	0.213	02:38
Breath Volume	1.687	02:38
Air Blank	0.000	02:38
Control Test	0.079	02:38
Air Blank	0.000	02:39
Diagnostics Check	OK	02:39

Data Download Date/Time: 10/02/2019 11:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,JONATAHN,R
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: SMITH,CAITLIN
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 08/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: BUSH

DATE/TIME OF TEST: 09/06/2019 02:50:33

Diagnostics Check	OK	02:44
Air Blank	0.000	02:44
Control Test	0.079	02:45
Air Blank	0.000	02:45
Subject Sample # 1	0.117	02:46
Breath Volume	1.902	02:46
Air Blank	0.000	02:46
Air Blank	0.000	02:48
Subject Sample # 2	0.121	02:49
Breath Volume	1.320	02:49
Air Blank	0.000	02:49
Control Test	0.079	02:49
Air Blank	0.000	02:50
Diagnostics Check	OK	02:50

Data Download Date/Time: 10/02/2019 11:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIS,DAVID,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: HARRIS,DAVID
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 08/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 09:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:30

SUBJECT'S LAST NAME: TICER

DATE/TIME OF TEST: 09/16/2019 09:57:44

Diagnostics Check	OK	09:51
Air Blank	0.000	09:51
Control Test	0.079	09:51
Air Blank	0.000	09:52
Subject Sample # 1	0.262	09:53
Breath Volume	1.183	09:53
Air Blank	0.000	09:53
Air Blank	0.000	09:55
Subject Sample # 2	0.259	09:56
Breath Volume	1.574	09:56
Air Blank	0.000	09:56
Control Test	0.078	09:57
Air Blank	0.000	09:57
Diagnostics Check	OK	09:57

Data Download Date/Time: 10/02/2019 11:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,OFC,J
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: ROSSOMONDO,CPL
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 08/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: BLOCKER

DATE/TIME OF TEST: 09/26/2019 22:47:19

Diagnostics Check	OK	22:41
Air Blank	0.000	22:41
Control Test	0.080	22:41
Air Blank	0.000	22:42
Subject Sample # 1	0.000	22:42
Breath Volume	3.824	22:42
Air Blank	0.000	22:43
Air Blank	0.000	22:45
Subject Sample # 2	0.000	22:45
Breath Volume	2.878	22:45
Air Blank	0.000	22:46
Control Test	0.080	22:46
Air Blank	0.000	22:47
Diagnostics Check	OK	22:47

Data Download Date/Time: 10/02/2019 11:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001207

INSTRUMENT REGISTERED TO: HENDRY COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORWOOD,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: WOOSLEY,LAWRENCE
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 08/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: CURRY

DATE/TIME OF TEST: 09/14/2019 02:10:12

Diagnostics Check	OK	02:03
Air Blank	0.000	02:04
Control Test	0.081	02:04
Air Blank	0.000	02:05
Subject Sample # 1	0.177	02:05
Breath Volume	1.367	02:05
Air Blank	0.000	02:06
Air Blank	0.000	02:07
Subject Sample # 2	0.178	02:08
Breath Volume	1.882	02:08
Air Blank	0.000	02:09
Control Test	0.082	02:09
Air Blank	0.000	02:10
Diagnostics Check	OK	02:10

Data Download Date/Time: 09/24/2019 01:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001207

INSTRUMENT REGISTERED TO: HENDRY COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORWOOD,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: WOOSLEY,LAWRENCE
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 09/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: VILLATORO

DATE/TIME OF TEST: 09/28/2019 01:01:41

Diagnostics Check	OK	00:58	
Air Blank	0.000	00:59	
Control Test	0.081	00:59	
Air Blank	0.000	01:00	
Subject Sample	# 1	01:00	Subject Test Refused
Breath Volume	0.000	01:00	
Air Blank	0.000	01:00	
Control Test	0.081	01:01	
Air Blank	0.000	01:01	
Diagnostics Check	OK	01:01	

Data Download Date/Time: 10/14/2019 10:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

INSTRUMENT REGISTERED TO: AUBURNDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAVRANEK, ADAM, J
BREATH TEST OPERATOR AGENCY: Auburndale PD
ARRESTING OFFICER: SEYMOUR, KEVIN
ARRESTING OFFICER AGENCY: Auburndale PD
LAST AGENCY INSPECTION: 08/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:37

SUBJECT'S LAST NAME: STARLING

DATE/TIME OF TEST: 09/03/2019 20:09:11

Diagnostics Check	OK	20:06	
Air Blank	0.000	20:06	
Control Test	0.080	20:07	
Air Blank	0.000	20:07	
Subject Sample	# 1	20:07	Subject Test Refused
Breath Volume	0.000	20:07	
Air Blank	0.000	20:08	
Control Test	0.080	20:08	
Air Blank	0.000	20:09	
Diagnostics Check	OK	20:09	

Data Download Date/Time: 09/18/2019 13:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

INSTRUMENT REGISTERED TO: AUBURNDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAVRANEK, ADAM, J
BREATH TEST OPERATOR AGENCY: Auburndale PD
ARRESTING OFFICER: LAHR, TRISTEN
ARRESTING OFFICER AGENCY: Auburndale PD
LAST AGENCY INSPECTION: 09/18/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: BANKS

DATE/TIME OF TEST: 09/24/2019 01:34:09

Diagnostics Check	OK	01:18	
Air Blank	0.000	01:18	
Control Test	0.080	01:18	
Air Blank	0.000	01:19	
Subject Sample	# 1 0.154	01:22	Volume Not Met
Breath Volume	0.292	01:22	
Air Blank	0.000	01:23	
Air Blank	0.000	01:25	
Subject Sample	# 2 0.168	01:26	
Breath Volume	1.261	01:26	
Air Blank	0.000	01:27	
Air Blank	0.000	01:29	
Subject Sample	# 3 0.144	01:32	Volume Not Met
Breath Volume	0.343	01:32	
Air Blank	0.000	01:33	
Control Test	0.077	01:33	
Air Blank	0.000	01:34	
Diagnostics Check	OK	01:34	

Data Download Date/Time: 10/15/2019 12:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DONDE,J,
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: TESTA,ANTHONY
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 08/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: FOULKS

DATE/TIME OF TEST: 09/14/2019 22:19:16

Diagnostics Check	OK	22:15	
Air Blank	0.000	22:16	
Control Test	0.081	22:16	
Air Blank	0.000	22:17	
Subject Sample	# 1	22:17	Subject Test Refused
Breath Volume	0.000	22:17	
Air Blank	0.000	22:18	
Control Test	0.081	22:18	
Air Blank	0.000	22:19	
Diagnostics Check	OK	22:19	

Data Download Date/Time: 09/30/2019 08:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DONDE,J,
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: LOPEZ,N
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 08/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:13

SUBJECT'S LAST NAME: WINN

DATE/TIME OF TEST: 09/22/2019 19:50:53

Diagnostics Check	OK	19:44
Air Blank	0.000	19:44
Control Test	0.080	19:44
Air Blank	0.000	19:45
Subject Sample # 1	0.210	19:46
Breath Volume	2.222	19:46
Air Blank	0.000	19:46
Air Blank	0.000	19:48
Subject Sample # 2	0.209	19:49
Breath Volume	2.269	19:49
Air Blank	0.000	19:49
Control Test	0.079	19:50
Air Blank	0.000	19:50
Diagnostics Check	OK	19:50

Data Download Date/Time: 09/30/2019 08:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,ROBERT,E
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HAMILTON,ROBERT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18018080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: VITERVO MONTEALEGRE

DATE/TIME OF TEST: 09/01/2019 01:14:42

Diagnostics Check	OK	01:07
Air Blank	0.000	01:07
Control Test	0.080	01:08
Air Blank	0.000	01:08
Subject Sample # 1	0.177	01:10
Breath Volume	1.414	01:10
Air Blank	0.000	01:10
Air Blank	0.000	01:12
Subject Sample # 2	0.174	01:13
Breath Volume	1.503	01:13
Air Blank	0.000	01:13
Control Test	0.079	01:14
Air Blank	0.000	01:14
Diagnostics Check	OK	01:14

Data Download Date/Time: 09/25/2019 06:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,ROBERT,E
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: CARTER,JACOB
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18018080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: TOMASELLO

DATE/TIME OF TEST: 09/01/2019 01:36:22

Diagnostics Check	OK	01:29
Air Blank	0.000	01:29
Control Test	0.080	01:29
Air Blank	0.000	01:30
Subject Sample # 1	0.000	01:31
Breath Volume	1.859	01:31
Air Blank	0.000	01:31
Air Blank	0.000	01:33
Subject Sample # 2	0.000	01:35
Breath Volume	2.648	01:35
Air Blank	0.000	01:35
Control Test	0.079	01:35
Air Blank	0.000	01:36
Diagnostics Check	OK	01:36

Data Download Date/Time: 09/25/2019 06:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES,STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: PROPST,JOHN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: ANCAYA LUJAN

DATE/TIME OF TEST: 09/01/2019 03:19:23

Diagnostics Check	OK	03:08	
Air Blank	0.000	03:08	
Control Test	0.079	03:08	
Air Blank	0.000	03:09	
Subject Sample	# 1	03:12	No Sample Provided
Breath Volume	0.000	03:12	
Air Blank	0.000	03:12	
Air Blank	0.000	03:14	
Subject Sample	# 2	03:18	No Sample Provided
Breath Volume	0.000	03:18	
Air Blank	0.000	03:18	
Control Test	0.079	03:18	
Air Blank	0.000	03:19	
Diagnostics Check	OK	03:19	

Data Download Date/Time: 09/25/2019 06:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,ROBERT,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: TURNER,LATORIAN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: ESCOBAR COLON

DATE/TIME OF TEST: 09/01/2019 03:30:27

Diagnostics Check	OK	03:20
Air Blank	0.000	03:21
Control Test	0.079	03:21
Air Blank	0.000	03:22
Subject Sample # 1	0.131	03:25
Breath Volume	1.238	03:25
Air Blank	0.000	03:25
Air Blank	0.000	03:27
Subject Sample # 2	0.137	03:28
Breath Volume	1.984	03:28
Air Blank	0.000	03:29
Control Test	0.079	03:29
Air Blank	0.000	03:30
Diagnostics Check	OK	03:30

Data Download Date/Time: 09/25/2019 06:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIGGINBOTHAM,JACK,C
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HIGGINBOTHAM,JACK
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18018080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: PEREZ MCDOUGAL

DATE/TIME OF TEST: 09/01/2019 03:51:23

Diagnostics Check	OK	03:44
Air Blank	0.000	03:45
Control Test	0.078	03:45
Air Blank	0.000	03:46
Subject Sample # 1	0.130	03:46
Breath Volume	2.660	03:46
Air Blank	0.000	03:47
Air Blank	0.000	03:49
Subject Sample # 2	0.147	03:49
Breath Volume	2.097	03:49
Air Blank	0.000	03:50
Control Test	0.078	03:50
Air Blank	0.000	03:51
Diagnostics Check	OK	03:51

Data Download Date/Time: 09/25/2019 06:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUCAS,NATHANIEL,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LUCAS,NATHANIEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: GOMEZ MORALES

DATE/TIME OF TEST: 09/01/2019 05:42:42

Diagnostics Check	OK	05:31
Air Blank	0.000	05:31
Control Test	0.079	05:31
Air Blank	0.000	05:32
Subject Sample # 1	0.215	05:33
Breath Volume	1.167	05:33
Air Blank	0.000	05:34
Air Blank	0.000	05:36
Subject Sample # 2	0.237	05:38
Breath Volume	1.289	05:38
Air Blank	0.000	05:38
Air Blank	0.000	05:40
Subject Sample # 3	0.236	05:41
Breath Volume	1.699	05:41
Air Blank	0.000	05:41
Control Test	0.078	05:42
Air Blank	0.000	05:42
Diagnostics Check	OK	05:42

Data Download Date/Time: 09/25/2019 06:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,ROBERT,E
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HAMILTON,ROBERT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18018080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:01

SUBJECT'S LAST NAME: WILDER

DATE/TIME OF TEST: 09/01/2019 16:34:29

Diagnostics Check	OK	16:28
Air Blank	0.000	16:29
Control Test	0.080	16:29
Air Blank	0.000	16:29
Subject Sample # 1	0.000	16:30
Breath Volume	2.500	16:30
Air Blank	0.000	16:30
Air Blank	0.000	16:32
Subject Sample # 2	0.000	16:33
Breath Volume	2.636	16:33
Air Blank	0.000	16:33
Control Test	0.078	16:33
Air Blank	0.000	16:34
Diagnostics Check	OK	16:34

Data Download Date/Time: 09/25/2019 06:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,ROBERT,E
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: PROPST,JOHN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18018080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: PASSLEY

DATE/TIME OF TEST: 09/01/2019 18:52:01

Diagnostics Check	OK	18:43	
Air Blank	0.000	18:44	
Control Test	0.080	18:44	
Air Blank	0.000	18:44	
Subject Sample	# 1	18:48	No Sample Provided
Breath Volume	0.000	18:48	
Air Blank	0.000	18:48	
Air Blank	0.000	18:50	
Subject Sample	# 2	18:50	Subject Test Refused
Breath Volume	0.000	18:50	
Air Blank	0.000	18:51	
Control Test	0.079	18:51	
Air Blank	0.000	18:51	
Diagnostics Check	OK	18:52	

Data Download Date/Time: 09/25/2019 06:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,ROBERT,E
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HAMILTON,ROBERT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18018080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 09/02/2019 00:48:15

Diagnostics Check	OK	00:42
Air Blank	0.000	00:42
Control Test	0.079	00:42
Air Blank	0.000	00:43
Subject Sample # 1	0.110	00:43
Breath Volume	1.617	00:43
Air Blank	0.000	00:44
Air Blank	0.000	00:46
Subject Sample # 2	0.111	00:46
Breath Volume	1.894	00:46
Air Blank	0.000	00:47
Control Test	0.079	00:47
Air Blank	0.000	00:48
Diagnostics Check	OK	00:48

Data Download Date/Time: 09/25/2019 06:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIGGINBOTHAM,JACK,C
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HIGGINBOTHAM,JACK
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18018080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 09/02/2019 01:15:15

Diagnostics Check	OK	01:08
Air Blank	0.000	01:09
Control Test	0.079	01:09
Air Blank	0.000	01:10
Subject Sample # 1	0.153	01:10
Breath Volume	1.191	01:10
Air Blank	0.000	01:11
Air Blank	0.000	01:13
Subject Sample # 2	0.169	01:13
Breath Volume	2.054	01:13
Air Blank	0.000	01:14
Control Test	0.078	01:14
Air Blank	0.000	01:15
Diagnostics Check	OK	01:15

Data Download Date/Time: 09/25/2019 06:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,DALE,E
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ROBERTS,DALE
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: LARREGUI-MAISONET

DATE/TIME OF TEST: 09/03/2019 04:44:49

Diagnostics Check	OK	04:37
Air Blank	0.000	04:38
Control Test	0.079	04:38
Air Blank	0.000	04:38
Subject Sample # 1	0.075	04:40
Breath Volume	1.476	04:40
Air Blank	0.000	04:41
Air Blank	0.000	04:42
Subject Sample # 2	0.072	04:43
Breath Volume	1.929	04:43
Air Blank	0.000	04:43
Control Test	0.077	04:44
Air Blank	0.000	04:44
Diagnostics Check	OK	04:44

Data Download Date/Time: 09/25/2019 06:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,DALE,E
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:23

SUBJECT'S LAST NAME: HARPER

DATE/TIME OF TEST: 09/03/2019 20:02:05

Diagnostics Check	OK	19:51	
Air Blank	0.000	19:52	
Control Test	0.079	19:52	
Air Blank	0.000	19:52	
Subject Sample	# 1 0.229	19:56	Volume Not Met
Breath Volume	0.699	19:56	
Air Blank	0.000	19:56	
Air Blank	0.000	19:58	
Subject Sample	# 2	20:00	Subject Test Refused
Breath Volume	0.699	20:00	
Air Blank	0.000	20:01	
Control Test	0.078	20:01	
Air Blank	0.000	20:02	
Diagnostics Check	OK	20:02	

Data Download Date/Time: 09/25/2019 06:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINET,RICKY,H
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BINET,RICKY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 11:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:52

SUBJECT'S LAST NAME: HUNT

DATE/TIME OF TEST: 09/04/2019 12:02:38

Diagnostics Check	OK	11:56
Air Blank	0.000	11:57
Control Test	0.079	11:57
Air Blank	0.000	11:57
Subject Sample # 1	0.000	11:58
Breath Volume	1.855	11:58
Air Blank	0.000	11:58
Air Blank	0.000	12:00
Subject Sample # 2	0.000	12:01
Breath Volume	3.433	12:01
Air Blank	0.000	12:01
Control Test	0.078	12:02
Air Blank	0.000	12:02
Diagnostics Check	OK	12:02

Data Download Date/Time: 09/25/2019 06:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHRISTELLO, CLIFFORD, A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LONDON, JOHN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: BARNETT

DATE/TIME OF TEST: 09/05/2019 03:26:32

Diagnostics Check	OK	03:23	
Air Blank	0.000	03:24	
Control Test	0.080	03:24	
Air Blank	0.000	03:24	
Subject Sample	# 1	03:25	Subject Test Refused
Breath Volume	0.000	03:25	
Air Blank	0.000	03:25	
Control Test	0.079	03:25	
Air Blank	0.000	03:26	
Diagnostics Check	OK	03:26	

Data Download Date/Time: 09/25/2019 06:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,ROBERT,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HAMILTON,ROBERT
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:29

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 09/05/2019 20:58:10

Diagnostics Check	OK	20:49	
Air Blank	0.000	20:50	
Control Test	0.080	20:50	
Air Blank	0.000	20:51	
Subject Sample	# 1	20:54	No Sample Provided
Breath Volume	0.000	20:54	
Air Blank	0.000	20:54	
Air Blank	0.000	20:56	
Subject Sample	# 2	20:56	Subject Test Refused
Breath Volume	0.000	20:56	
Air Blank	0.000	20:57	
Control Test	0.079	20:57	
Air Blank	0.000	20:58	
Diagnostics Check	OK	20:58	

Data Download Date/Time: 09/25/2019 06:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAUGHERTY,KASIE,L
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HAMPTON,JAMES
ARRESTING OFFICER AGENCY: Lake Wales PD
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: REED

DATE/TIME OF TEST: 09/05/2019 23:09:24

Diagnostics Check	OK	23:02
Air Blank	0.000	23:02
Control Test	0.080	23:03
Air Blank	0.000	23:03
Subject Sample # 1	0.000	23:05
Breath Volume	2.765	23:05
Air Blank	0.000	23:05
Air Blank	0.000	23:07
Subject Sample # 2	0.000	23:08
Breath Volume	2.105	23:08
Air Blank	0.000	23:08
Control Test	0.079	23:08
Air Blank	0.000	23:09
Diagnostics Check	OK	23:09

Data Download Date/Time: 09/25/2019 06:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELTON,X,V
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: TAYLOR,G
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: HAIN

DATE/TIME OF TEST: 09/06/2019 23:15:17

Diagnostics Check	OK	23:12	
Air Blank	0.000	23:12	
Control Test	0.079	23:12	
Air Blank	0.000	23:13	
Subject Sample	# 1	23:13	Subject Test Refused
Breath Volume	0.000	23:13	
Air Blank	0.000	23:14	
Control Test	0.079	23:14	
Air Blank	0.000	23:15	
Diagnostics Check	OK	23:15	

Data Download Date/Time: 09/25/2019 06:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINET,RICKY,H
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BINET,RICKY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:26

SUBJECT'S LAST NAME: FIGUEROA

DATE/TIME OF TEST: 09/07/2019 22:51:48

Diagnostics Check	OK	22:44
Air Blank	0.000	22:45
Control Test	0.079	22:45
Air Blank	0.000	22:46
Subject Sample # 1	0.230	22:47
Breath Volume	1.312	22:47
Air Blank	0.000	22:47
Air Blank	0.000	22:49
Subject Sample # 2	0.230	22:50
Breath Volume	1.175	22:50
Air Blank	0.000	22:50
Control Test	0.077	22:51
Air Blank	0.000	22:51
Diagnostics Check	OK	22:51

Data Download Date/Time: 09/25/2019 06:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,DALE,E
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: FLORES-MARTINEZ

DATE/TIME OF TEST: 09/08/2019 02:19:08

Diagnostics Check	OK	02:10
Air Blank	0.000	02:11
Control Test	0.080	02:11
Air Blank	0.000	02:11
Subject Sample # 1	0.169	02:13
Breath Volume	1.804	02:13
Air Blank	0.000	02:14
Air Blank	0.000	02:16
Subject Sample # 2	0.166	02:17
Breath Volume	1.472	02:17
Air Blank	0.000	02:18
Control Test	0.078	02:18
Air Blank	0.000	02:19
Diagnostics Check	OK	02:19

Data Download Date/Time: 09/25/2019 06:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,ROBERT,E
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HAMILTON,ROBERT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18018080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: STASIK

DATE/TIME OF TEST: 09/11/2019 02:26:10

Diagnostics Check	OK	02:17	
Air Blank	0.000	02:18	
Control Test	0.080	02:18	
Air Blank	0.000	02:19	
Subject Sample	# 1	02:22	No Sample Provided
Breath Volume	0.000	02:22	
Air Blank	0.000	02:22	
Air Blank	0.000	02:24	
Subject Sample	# 2	02:24	Subject Test Refused
Breath Volume	0.000	02:24	
Air Blank	0.000	02:25	
Control Test	0.079	02:25	
Air Blank	0.000	02:26	
Diagnostics Check	OK	02:26	

Data Download Date/Time: 09/25/2019 06:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELTON,XAVIER,V
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: REGISTER,JUSTIN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: CAPERS

DATE/TIME OF TEST: 09/12/2019 03:30:15

Diagnostics Check	OK	03:23
Air Blank	0.000	03:23
Control Test	0.079	03:23
Air Blank	0.000	03:24
Subject Sample # 1	0.149	03:24
Breath Volume	1.808	03:24
Air Blank	0.000	03:25
Air Blank	0.000	03:27
Subject Sample # 2	0.132	03:28
Breath Volume	1.449	03:28
Air Blank	0.000	03:29
Control Test	0.079	03:29
Air Blank	0.000	03:30
Diagnostics Check	OK	03:30

Data Download Date/Time: 09/25/2019 06:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER, ADAM, H
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: PROPST, JOHN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: RUSSELL

DATE/TIME OF TEST: 09/13/2019 04:02:36

Diagnostics Check	OK	03:59	
Air Blank	0.000	04:00	
Control Test	0.079	04:00	
Air Blank	0.000	04:01	
Subject Sample	# 1	04:01	Subject Test Refused
Breath Volume	0.000	04:01	
Air Blank	0.000	04:01	
Control Test	0.079	04:02	
Air Blank	0.000	04:02	
Diagnostics Check	OK	04:02	

Data Download Date/Time: 09/25/2019 06:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,ROBERT,E
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HAMILTON,ROBERT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18018080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: VAILLANT

DATE/TIME OF TEST: 09/13/2019 23:55:57

Diagnostics Check	OK	23:49
Air Blank	0.000	23:50
Control Test	0.079	23:50
Air Blank	0.000	23:50
Subject Sample # 1	0.102	23:51
Breath Volume	3.257	23:51
Air Blank	0.000	23:52
Air Blank	0.000	23:54
Subject Sample # 2	0.089	23:54
Breath Volume	2.812	23:54
Air Blank	0.000	23:55
Control Test	0.079	23:55
Air Blank	0.000	23:55
Diagnostics Check	OK	23:55

Data Download Date/Time: 09/25/2019 06:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES, STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SYKES, STEVEN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: FRAGOSO PICAZO

DATE/TIME OF TEST: 09/14/2019 01:15:10

Diagnostics Check	OK	01:08
Air Blank	0.000	01:09
Control Test	0.079	01:09
Air Blank	0.000	01:10
Subject Sample # 1	0.242	01:10
Breath Volume	2.187	01:10
Air Blank	0.000	01:11
Air Blank	0.000	01:13
Subject Sample # 2	0.231	01:13
Breath Volume	1.945	01:13
Air Blank	0.000	01:14
Control Test	0.079	01:14
Air Blank	0.000	01:15
Diagnostics Check	OK	01:15

Data Download Date/Time: 09/25/2019 06:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,ROBERT,E
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HAMILTON,ROBERT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18018080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: MITTON

DATE/TIME OF TEST: 09/14/2019 03:20:02

Diagnostics Check	OK	03:13
Air Blank	0.000	03:14
Control Test	0.079	03:14
Air Blank	0.000	03:14
Subject Sample # 1	0.126	03:15
Breath Volume	1.980	03:15
Air Blank	0.000	03:16
Air Blank	0.000	03:18
Subject Sample # 2	0.132	03:18
Breath Volume	2.300	03:18
Air Blank	0.000	03:19
Control Test	0.079	03:19
Air Blank	0.000	03:19
Diagnostics Check	OK	03:20

Data Download Date/Time: 09/25/2019 06:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,ROBERT,E
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HAMILTON,ROBERT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18018080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: THACKER

DATE/TIME OF TEST: 09/15/2019 03:19:55

Diagnostics Check	OK	03:11	
Air Blank	0.000	03:11	
Control Test	0.080	03:12	
Air Blank	0.000	03:12	
Subject Sample	# 1	03:15	No Sample Provided
Breath Volume	0.000	03:15	
Air Blank	0.000	03:16	
Air Blank	0.000	03:18	
Subject Sample	# 2	03:18	Subject Test Refused
Breath Volume	0.000	03:18	
Air Blank	0.000	03:19	
Control Test	0.080	03:19	
Air Blank	0.000	03:19	
Diagnostics Check	OK	03:19	

Data Download Date/Time: 09/25/2019 06:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES,STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SYKES,STEVEN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: ROSA RIVERA

DATE/TIME OF TEST: 09/15/2019 03:34:49

Diagnostics Check	OK	03:23
Air Blank	0.000	03:23
Control Test	0.079	03:24
Air Blank	0.000	03:24
Subject Sample # 1	0.190	03:27
Breath Volume	1.367	03:27
Air Blank	0.000	03:28
Air Blank	0.000	03:30
Subject Sample # 2	0.180	03:33
Breath Volume	1.363	03:33
Air Blank	0.000	03:33
Control Test	0.077	03:34
Air Blank	0.000	03:34
Diagnostics Check	OK	03:34

Data Download Date/Time: 09/25/2019 06:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,ROBERT,E
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HAMILTON,ROBERT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18018080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:09

SUBJECT'S LAST NAME: BENITO TOLENTINO

DATE/TIME OF TEST: 09/15/2019 20:43:32

Diagnostics Check	OK	20:36
Air Blank	0.000	20:36
Control Test	0.079	20:37
Air Blank	0.000	20:37
Subject Sample # 1	0.200	20:39
Breath Volume	1.925	20:39
Air Blank	0.000	20:39
Air Blank	0.000	20:41
Subject Sample # 2	0.183	20:41
Breath Volume	1.527	20:41
Air Blank	0.000	20:42
Control Test	0.079	20:42
Air Blank	0.000	20:43
Diagnostics Check	OK	20:43

Data Download Date/Time: 09/25/2019 06:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORAN-HERNANDEZ,JOSE,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HIGGINBOTHAM,JACK
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18018080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: MORAN-HERNANDEZ

DATE/TIME OF TEST: 09/15/2019 23:36:37

Diagnostics Check	OK	23:29
Air Blank	0.000	23:29
Control Test	0.079	23:29
Air Blank	0.000	23:30
Subject Sample # 1	0.145	23:31
Breath Volume	2.136	23:31
Air Blank	0.000	23:32
Air Blank	0.000	23:34
Subject Sample # 2	0.137	23:35
Breath Volume	2.160	23:35
Air Blank	0.000	23:35
Control Test	0.075	23:36
Air Blank	0.000	23:36
Diagnostics Check	OK	23:36

Data Download Date/Time: 09/25/2019 06:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIGGINBOTHAM,JACK,C
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HIGGINBOTHAM,JACK
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18018080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 09/20/2019 00:12:21

Diagnostics Check	OK	00:04
Air Blank	0.000	00:04
Control Test	0.079	00:05
Air Blank	0.000	00:05
Subject Sample # 1	0.101	00:07
Breath Volume	2.187	00:07
Air Blank	0.000	00:08
Air Blank	0.000	00:10
Subject Sample # 2	0.096	00:10
Breath Volume	2.281	00:10
Air Blank	0.000	00:11
Control Test	0.076	00:11
Air Blank	0.000	00:12
Diagnostics Check	OK	00:12

Data Download Date/Time: 09/25/2019 06:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES,STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SYKES,STEVEN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: CRAWFORD

DATE/TIME OF TEST: 09/20/2019 02:40:42

Diagnostics Check	OK	02:37	
Air Blank	0.000	02:38	
Control Test	0.080	02:38	
Air Blank	0.000	02:39	
Subject Sample	# 1	02:39	Subject Test Refused
Breath Volume	0.000	02:39	
Air Blank	0.000	02:39	
Control Test	0.078	02:40	
Air Blank	0.000	02:40	
Diagnostics Check	OK	02:40	

Data Download Date/Time: 09/25/2019 06:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES,STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SYKES,STEVEN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: CRISWELL

DATE/TIME OF TEST: 09/20/2019 04:23:18

Diagnostics Check	OK	04:17
Air Blank	0.000	04:17
Control Test	0.080	04:18
Air Blank	0.000	04:18
Subject Sample # 1	0.100	04:18
Breath Volume	1.289	04:18
Air Blank	0.000	04:19
Air Blank	0.000	04:21
Subject Sample # 2	0.101	04:21
Breath Volume	1.125	04:21
Air Blank	0.000	04:22
Control Test	0.079	04:22
Air Blank	0.000	04:23
Diagnostics Check	OK	04:23

Data Download Date/Time: 09/25/2019 06:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,DALE,E
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ROBERTS,DALE
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: LACY

DATE/TIME OF TEST: 09/21/2019 00:00:25

Diagnostics Check	OK	23:53
Air Blank	0.000	23:53
Control Test	0.079	23:53
Air Blank	0.000	23:54
Subject Sample # 1	0.223	23:54
Breath Volume	1.292	23:54
Air Blank	0.000	23:55
Air Blank	0.000	23:57
Subject Sample # 2	0.216	23:58
Breath Volume	2.312	23:58
Air Blank	0.000	23:59
Control Test	0.079	23:59
Air Blank	0.000	00:00
Diagnostics Check	OK	00:00

Data Download Date/Time: 09/25/2019 06:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELTON,XAVIER,V
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: REGISTER,JUSTIN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: ROGELIO

DATE/TIME OF TEST: 09/21/2019 04:20:30

Diagnostics Check	OK	04:13
Air Blank	0.000	04:14
Control Test	0.079	04:14
Air Blank	0.000	04:15
Subject Sample # 1	0.229	04:16
Breath Volume	2.859	04:16
Air Blank	0.000	04:16
Air Blank	0.000	04:18
Subject Sample # 2	0.230	04:18
Breath Volume	2.609	04:18
Air Blank	0.000	04:19
Control Test	0.077	04:19
Air Blank	0.000	04:20
Diagnostics Check	OK	04:20

Data Download Date/Time: 09/25/2019 06:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELCHER,ALEX,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: PROPST,JOHN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: MILLS

DATE/TIME OF TEST: 09/21/2019 23:55:11

Diagnostics Check	OK	23:48
Air Blank	0.000	23:49
Control Test	0.080	23:49
Air Blank	0.000	23:50
Subject Sample # 1	0.000	23:50
Breath Volume	3.343	23:50
Air Blank	0.000	23:51
Air Blank	0.000	23:53
Subject Sample # 2	0.000	23:53
Breath Volume	4.085	23:53
Air Blank	0.000	23:54
Control Test	0.079	23:54
Air Blank	0.000	23:55
Diagnostics Check	OK	23:55

Data Download Date/Time: 09/25/2019 06:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,DALE,E
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ROBERTS,DALE
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:10

SUBJECT'S LAST NAME: HERNANDEZ-ESCALANTE

DATE/TIME OF TEST: 09/22/2019 16:44:05

Diagnostics Check	OK	16:37
Air Blank	0.000	16:38
Control Test	0.079	16:38
Air Blank	0.000	16:39
Subject Sample # 1	0.178	16:39
Breath Volume	1.988	16:39
Air Blank	0.000	16:40
Air Blank	0.000	16:42
Subject Sample # 2	0.177	16:42
Breath Volume	2.417	16:42
Air Blank	0.000	16:43
Control Test	0.079	16:43
Air Blank	0.000	16:43
Diagnostics Check	OK	16:44

Data Download Date/Time: 09/25/2019 06:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JEUDY,RINALDO,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: CARTER,JACOB
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: BUI
ARREST TIME: 22:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: GARDNER

DATE/TIME OF TEST: 09/23/2019 00:32:57

Diagnostics Check	OK	00:30	
Air Blank	0.000	00:30	
Control Test	0.079	00:30	
Air Blank	0.000	00:31	
Subject Sample	# 1	00:31	Subject Test Refused
Breath Volume	0.000	00:31	
Air Blank	0.000	00:32	
Control Test	0.079	00:32	
Air Blank	0.000	00:32	
Diagnostics Check	OK	00:32	

Data Download Date/Time: 09/25/2019 06:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JEUDY,RINALDO,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JACKSON,JEFFREY
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: BUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: SCARBOROUGH

DATE/TIME OF TEST: 09/23/2019 00:40:42

Diagnostics Check	OK	00:37	
Air Blank	0.000	00:37	
Control Test	0.079	00:38	
Air Blank	0.000	00:38	
Subject Sample	# 1	00:39	Subject Test Refused
Breath Volume	0.000	00:39	
Air Blank	0.000	00:39	
Control Test	0.079	00:40	
Air Blank	0.000	00:40	
Diagnostics Check	OK	00:40	

Data Download Date/Time: 09/25/2019 06:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINET,RICKY,H
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BINET,RICKY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:15

SUBJECT'S LAST NAME: UNDERWOOD

DATE/TIME OF TEST: 09/23/2019 16:44:14

Diagnostics Check	OK	16:37
Air Blank	0.000	16:38
Control Test	0.080	16:38
Air Blank	0.000	16:39
Subject Sample # 1	0.080	16:39
Breath Volume	1.125	16:39
Air Blank	0.000	16:40
Air Blank	0.000	16:41
Subject Sample # 2	0.081	16:42
Breath Volume	2.406	16:42
Air Blank	0.000	16:43
Control Test	0.078	16:43
Air Blank	0.000	16:44
Diagnostics Check	OK	16:44

Data Download Date/Time: 09/25/2019 06:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIGGINBOTHAM,JACK,C
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HIGGINBOTHAM,JACK
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 09/23/2019 23:10:48

Diagnostics Check	OK	22:59	
Air Blank	0.000	22:59	
Control Test	0.080	22:59	
Air Blank	0.000	23:00	
Subject Sample	# 1 0.103	23:03	Volume Not Met
Breath Volume	0.273	23:03	
Air Blank	0.000	23:04	
Air Blank	0.000	23:06	
Subject Sample	# 2 0.099	23:09	Volume Not Met
Breath Volume	1.000	23:09	
Air Blank	0.000	23:09	
Control Test	0.078	23:10	
Air Blank	0.000	23:10	
Diagnostics Check	OK	23:10	

Data Download Date/Time: 09/25/2019 06:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,ROBERT,E
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HAMILTON,ROBERT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: FLORES

DATE/TIME OF TEST: 09/24/2019 01:05:20

Diagnostics Check	OK	00:59
Air Blank	0.000	00:59
Control Test	0.079	01:00
Air Blank	0.000	01:00
Subject Sample # 1	0.000	01:01
Breath Volume	3.703	01:01
Air Blank	0.000	01:01
Air Blank	0.000	01:03
Subject Sample # 2	0.000	01:03
Breath Volume	2.152	01:03
Air Blank	0.000	01:04
Control Test	0.079	01:04
Air Blank	0.000	01:05
Diagnostics Check	OK	01:05

Data Download Date/Time: 09/25/2019 06:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHOUINARD,JEFFREY,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: CHOUINARD,JEFFREY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: MANN

DATE/TIME OF TEST: 09/24/2019 03:14:02

Diagnostics Check	OK	03:11	
Air Blank	0.000	03:11	
Control Test	0.080	03:12	
Air Blank	0.000	03:12	
Subject Sample	# 1	03:12	Subject Test Refused
Breath Volume	0.000	03:12	
Air Blank	0.000	03:13	
Control Test	0.079	03:13	
Air Blank	0.000	03:13	
Diagnostics Check	OK	03:14	

Data Download Date/Time: 09/25/2019 06:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,ROBERT,E
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HAMILTON,ROBERT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:04

SUBJECT'S LAST NAME: BANEGAS

DATE/TIME OF TEST: 09/24/2019 20:38:29

Diagnostics Check	OK	20:31
Air Blank	0.000	20:32
Control Test	0.079	20:32
Air Blank	0.000	20:33
Subject Sample # 1	0.169	20:33
Breath Volume	1.964	20:33
Air Blank	0.000	20:34
Air Blank	0.000	20:36
Subject Sample # 2	0.162	20:36
Breath Volume	1.785	20:36
Air Blank	0.000	20:37
Control Test	0.079	20:37
Air Blank	0.000	20:38
Diagnostics Check	OK	20:38

Data Download Date/Time: 09/25/2019 06:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIGGINBOTHAM,JACK,C
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HIGGINBOTHAM,JACK
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: GRASSI

DATE/TIME OF TEST: 09/25/2019 01:43:02

Diagnostics Check	OK	01:31	
Air Blank	0.000	01:32	
Control Test	0.080	01:32	
Air Blank	0.000	01:32	
Subject Sample	# 1	01:36	No Sample Provided
Breath Volume	0.000	01:36	
Air Blank	0.000	01:36	
Air Blank	0.000	01:38	
Subject Sample	# 2	01:41	No Sample Provided
Breath Volume	0.000	01:41	
Air Blank	0.000	01:42	
Control Test	0.079	01:42	
Air Blank	0.000	01:42	
Diagnostics Check	OK	01:43	

Data Download Date/Time: 09/25/2019 06:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIGGINBOTHAM,JACK,C
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HIGGINBOTHAM,JACK
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: GRASSI

DATE/TIME OF TEST: 09/25/2019 02:04:00

Diagnostics Check	OK	01:47	
Air Blank	0.000	01:48	
Control Test	0.079	01:48	
Air Blank	0.000	01:48	
Subject Sample	# 1 0.192	01:52	Volume Not Met
Breath Volume	0.035	01:52	
Air Blank	0.000	01:52	
Air Blank	0.000	01:54	
Subject Sample	# 2 0.224	01:56	
Breath Volume	1.933	01:56	
Air Blank	0.000	01:57	
Air Blank	0.000	01:58	
Subject Sample	# 3 0.172	02:02	Volume Not Met
Breath Volume	0.644	02:02	
Air Blank	0.000	02:03	
Control Test	0.074	02:03	Control Outside Tolerance
Air Blank	0.000	02:04	

Data Download Date/Time: 09/25/2019 06:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,DALE,E
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: PROPST,JOHN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 09/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 09/26/2019 00:08:27

Diagnostics Check	OK	00:01
Air Blank	0.000	00:01
Control Test	0.079	00:02
Air Blank	0.000	00:02
Subject Sample # 1	0.135	00:03
Breath Volume	2.710	00:03
Air Blank	0.000	00:04
Air Blank	0.000	00:06
Subject Sample # 2	0.138	00:06
Breath Volume	2.125	00:06
Air Blank	0.000	00:07
Control Test	0.079	00:07
Air Blank	0.000	00:08
Diagnostics Check	OK	00:08

Data Download Date/Time: 10/23/2019 09:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELANEY,AMY,L
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: TURNAGE,M
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 09/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 12:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:33

SUBJECT'S LAST NAME: WEARING

DATE/TIME OF TEST: 09/26/2019 13:17:04

Diagnostics Check	OK	13:09
Air Blank	0.000	13:10
Control Test	0.079	13:10
Air Blank	0.000	13:10
Subject Sample # 1	0.161	13:12
Breath Volume	2.820	13:12
Air Blank	0.000	13:13
Air Blank	0.000	13:15
Subject Sample # 2	0.160	13:15
Breath Volume	2.042	13:15
Air Blank	0.000	13:16
Control Test	0.078	13:16
Air Blank	0.000	13:16
Diagnostics Check	OK	13:17

Data Download Date/Time: 10/23/2019 09:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,DALE,E
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LONDON,JOHN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 09/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: FONSECA

DATE/TIME OF TEST: 09/27/2019 03:39:00

Diagnostics Check	OK	03:35	
Air Blank	0.000	03:36	
Control Test	0.080	03:36	
Air Blank	0.000	03:37	
Subject Sample	# 1	03:37	Subject Test Refused
Breath Volume	0.000	03:37	
Air Blank	0.000	03:38	
Control Test	0.079	03:38	
Air Blank	0.000	03:38	
Diagnostics Check	OK	03:39	

Data Download Date/Time: 10/23/2019 09:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,ROBERT,E
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HAMILTON,ROBERT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 09/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:56

SUBJECT'S LAST NAME: FREEMAN

DATE/TIME OF TEST: 09/28/2019 19:27:30

Diagnostics Check	OK	19:19
Air Blank	0.000	19:20
Control Test	0.079	19:20
Air Blank	0.000	19:21
Subject Sample # 1	0.272	19:23
Breath Volume	1.726	19:23
Air Blank	0.000	19:23
Air Blank	0.000	19:25
Subject Sample # 2	0.263	19:25
Breath Volume	1.156	19:25
Air Blank	0.000	19:26
Control Test	0.077	19:26
Air Blank	0.000	19:27
Diagnostics Check	OK	19:27

Data Download Date/Time: 10/23/2019 09:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIGGINBOTHAM,JACK,C
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HIGGINBOTHAM,JACK
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 09/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:22

SUBJECT'S LAST NAME: DRAWHORN

DATE/TIME OF TEST: 09/28/2019 21:08:10

Diagnostics Check	OK	20:58
Air Blank	0.000	20:58
Control Test	0.078	20:59
Air Blank	0.000	20:59
Subject Sample # 1	0.226	21:00
Breath Volume	1.753	21:00
Air Blank	0.000	21:01
Air Blank	0.000	21:03
Subject Sample # 2	0.211	21:06
Breath Volume	2.320	21:06
Air Blank	0.000	21:07
Control Test	0.078	21:07
Air Blank	0.000	21:08
Diagnostics Check	OK	21:08

Data Download Date/Time: 10/23/2019 09:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,ROBERT,E
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HAMILTON,ROBERT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 09/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: MITCHELL JR

DATE/TIME OF TEST: 09/29/2019 02:27:49

Diagnostics Check	OK	02:20
Air Blank	0.000	02:21
Control Test	0.078	02:21
Air Blank	0.000	02:21
Subject Sample # 1	0.163	02:22
Breath Volume	1.285	02:22
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample # 2	0.172	02:26
Breath Volume	1.558	02:26
Air Blank	0.000	02:26
Control Test	0.078	02:27
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

Data Download Date/Time: 10/23/2019 09:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: APONTE, JOSHUA,
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: APONTE, JOSHUA
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 09/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: SHAVER

DATE/TIME OF TEST: 09/26/2019 23:42:30

Diagnostics Check	OK	23:39	
Air Blank	0.000	23:40	
Control Test	0.082	23:40	
Air Blank	0.000	23:40	
Subject Sample	# 1	23:41	Subject Test Refused
Breath Volume	0.000	23:41	
Air Blank	0.000	23:41	
Control Test	0.081	23:41	
Air Blank	0.000	23:42	
Diagnostics Check	OK	23:42	

Data Download Date/Time: 10/24/2019 00:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUMBLISS, JEREMY, F
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: BENTIL, GERALD
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: CRUMBLISS

DATE/TIME OF TEST: 09/27/2019 03:22:45

Diagnostics Check	OK	03:19	
Air Blank	0.000	03:20	
Control Test	0.081	03:20	
Air Blank	0.000	03:20	
Subject Sample	# 1	03:21	Subject Test Refused
Breath Volume	0.000	03:21	
Air Blank	0.000	03:21	
Control Test	0.080	03:22	
Air Blank	0.000	03:22	
Diagnostics Check	OK	03:22	

Data Download Date/Time: 10/24/2019 00:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALSH,SHANNON,L
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: BROCATO,J
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 09/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: GARRARD

DATE/TIME OF TEST: 09/28/2019 02:49:55

Diagnostics Check	OK	02:42
Air Blank	0.000	02:42
Control Test	0.081	02:43
Air Blank	0.000	02:43
Subject Sample # 1	0.179	02:45
Breath Volume	3.152	02:45
Air Blank	0.000	02:45
Air Blank	0.000	02:47
Subject Sample # 2	0.182	02:48
Breath Volume	2.921	02:48
Air Blank	0.000	02:49
Control Test	0.081	02:49
Air Blank	0.000	02:49
Diagnostics Check	OK	02:49

Data Download Date/Time: 10/24/2019 00:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHESSHER,SANDRA,W
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: MURHCINSON,R
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 09/25/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:57

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 09/28/2019 21:20:35

Diagnostics Check	OK	21:17	
Air Blank	0.000	21:17	
Control Test	0.082	21:18	
Air Blank	0.000	21:18	
Subject Sample	# 1	21:19	Subject Test Refused
Breath Volume	0.000	21:19	
Air Blank	0.000	21:19	
Control Test	0.081	21:19	
Air Blank	0.000	21:20	
Diagnostics Check	OK	21:20	

Data Download Date/Time: 10/24/2019 00:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,GREGORY,A
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: REED,GREGORY
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: GRISSOM

DATE/TIME OF TEST: 09/30/2019 23:27:44

Diagnostics Check	OK	23:24	
Air Blank	0.000	23:24	
Control Test	0.082	23:25	
Air Blank	0.000	23:25	
Subject Sample	# 1	23:26	Subject Test Refused
Breath Volume	0.000	23:26	
Air Blank	0.000	23:26	
Control Test	0.081	23:27	
Air Blank	0.000	23:27	
Diagnostics Check	OK	23:27	

Data Download Date/Time: 10/24/2019 00:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRESENZI,G,
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: CRESENZI,G
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 08/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 765462
DRY GAS EXPIRY DATE: 04/13/2020
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: HOBBS

DATE/TIME OF TEST: 09/05/2019 01:26:14

Diagnostics Check	OK	01:18
Air Blank	0.000	01:19
Control Test	0.075	01:19
Air Blank	0.000	01:20
Subject Sample # 1	0.039	01:21
Breath Volume	1.636	01:21
Air Blank	0.000	01:22
Air Blank	0.000	01:23
Subject Sample # 2	0.040	01:24
Breath Volume	1.867	01:24
Air Blank	0.000	01:25
Control Test	0.075	01:25
Air Blank	0.000	01:26
Diagnostics Check	OK	01:26

Data Download Date/Time: 09/26/2019 09:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SLOAN,JULIAN,E
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: POTTER,KYLE
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 08/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 765462
DRY GAS EXPIRY DATE: 04/13/2020
VIOLATION CODE: DUI
ARREST TIME: 11:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:18

SUBJECT'S LAST NAME: MARTION

DATE/TIME OF TEST: 09/05/2019 11:51:47

Diagnostics Check	OK	11:40	
Air Blank	0.000	11:40	
Control Test	0.075	11:41	
Air Blank	0.000	11:41	
Subject Sample	# 1	11:44	No Sample Provided
Breath Volume	0.000	11:44	
Air Blank	0.000	11:45	
Air Blank	0.000	11:47	
Subject Sample	# 2	11:50	No Sample Provided
Breath Volume	0.000	11:50	
Air Blank	0.000	11:50	
Control Test	0.075	11:51	
Air Blank	0.000	11:51	
Diagnostics Check	OK	11:51	

Data Download Date/Time: 09/26/2019 09:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINZ,ANTHONY,E
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: CHRISTIE,CARLEY
ARRESTING OFFICER AGENCY: OTHER Fed
LAST AGENCY INSPECTION: 08/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 765462
DRY GAS EXPIRY DATE: 04/13/2020
VIOLATION CODE: DUI
ARREST TIME: 17:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: ABBOTT

DATE/TIME OF TEST: 09/25/2019 17:58:04

Diagnostics Check	OK	17:50
Air Blank	0.000	17:50
Control Test	0.075	17:51
Air Blank	0.000	17:51
Subject Sample # 1	0.121	17:53
Breath Volume	2.726	17:53
Air Blank	0.000	17:54
Air Blank	0.000	17:55
Subject Sample # 2	0.123	17:56
Breath Volume	1.628	17:56
Air Blank	0.000	17:57
Control Test	0.075	17:57
Air Blank	0.000	17:57
Diagnostics Check	OK	17:58

Data Download Date/Time: 09/26/2019 09:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001244

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIMA,MICHAEL,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: LIMA,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A3
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: STEEDLEY

DATE/TIME OF TEST: 09/10/2019 05:16:37

Diagnostics Check	OK	05:10
Air Blank	0.000	05:10
Control Test	0.080	05:11
Air Blank	0.000	05:11
Subject Sample # 1	0.131	05:12
Breath Volume	1.671	05:12
Air Blank	0.000	05:12
Air Blank	0.000	05:14
Subject Sample # 2	0.130	05:15
Breath Volume	2.582	05:15
Air Blank	0.000	05:15
Control Test	0.080	05:16
Air Blank	0.000	05:16
Diagnostics Check	OK	05:16

Data Download Date/Time: 10/05/2019 19:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOTZ,KATHRYN,N
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: GOTZ,KATHRYN
ARRESTING OFFICER AGENCY: Daytona Beach Shores
LAST AGENCY INSPECTION: 09/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1185592
DRY GAS EXPIRY DATE: 08/13/2022
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: THORNTON

DATE/TIME OF TEST: 09/23/2019 04:12:56

Diagnostics Check	OK	04:06
Air Blank	0.000	04:07
Control Test	0.080	04:07
Air Blank	0.000	04:08
Subject Sample # 1	0.158	04:08
Breath Volume	3.015	04:08
Air Blank	0.000	04:09
Air Blank	0.000	04:10
Subject Sample # 2	0.156	04:11
Breath Volume	2.750	04:11
Air Blank	0.000	04:12
Control Test	0.081	04:12
Air Blank	0.000	04:12
Diagnostics Check	OK	04:12

Data Download Date/Time: 10/21/2019 03:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIEDESCH, LAURA, R
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: LONG, JESSICA
ARRESTING OFFICER AGENCY: Daytona Beach Shores
LAST AGENCY INSPECTION: 09/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1185592
DRY GAS EXPIRY DATE: 08/13/2022
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 09/26/2019 04:49:28

Diagnostics Check	OK	04:43
Air Blank	0.000	04:43
Control Test	0.081	04:44
Air Blank	0.000	04:44
Subject Sample # 1	0.239	04:45
Breath Volume	1.726	04:45
Air Blank	0.000	04:45
Air Blank	0.000	04:47
Subject Sample # 2	0.232	04:47
Breath Volume	1.875	04:47
Air Blank	0.000	04:48
Control Test	0.081	04:48
Air Blank	0.000	04:49
Diagnostics Check	OK	04:49

Data Download Date/Time: 10/21/2019 03:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHNIBBEN,BRIAN,K
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: SCHNIBBEN,BRIAN
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 09/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1185592
DRY GAS EXPIRY DATE: 08/13/2022
VIOLATION CODE: DUI
ARREST TIME: 08:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:39

SUBJECT'S LAST NAME: MOHAMED

DATE/TIME OF TEST: 09/28/2019 09:10:14

Diagnostics Check	OK	09:02
Air Blank	0.000	09:03
Control Test	0.081	09:03
Air Blank	0.000	09:04
Subject Sample # 1	0.107	09:05
Breath Volume	4.253	09:05
Air Blank	0.000	09:06
Air Blank	0.000	09:07
Subject Sample # 2	0.110	09:08
Breath Volume	1.234	09:08
Air Blank	0.000	09:09
Control Test	0.081	09:09
Air Blank	0.000	09:10
Diagnostics Check	OK	09:10

Data Download Date/Time: 10/21/2019 03:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMBSH,CARLYTON,R
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: HAMBSH,CARLYTON
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 09/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1185592
DRY GAS EXPIRY DATE: 08/13/2022
VIOLATION CODE: DUI
ARREST TIME: 18:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:57

SUBJECT'S LAST NAME: KYER

DATE/TIME OF TEST: 09/28/2019 19:30:27

Diagnostics Check	OK	19:23
Air Blank	0.000	19:24
Control Test	0.082	19:24
Air Blank	0.000	19:25
Subject Sample # 1	0.093	19:25
Breath Volume	3.019	19:25
Air Blank	0.000	19:26
Air Blank	0.000	19:28
Subject Sample # 2	0.096	19:28
Breath Volume	2.773	19:28
Air Blank	0.000	19:29
Control Test	0.081	19:29
Air Blank	0.000	19:30
Diagnostics Check	OK	19:30

Data Download Date/Time: 10/21/2019 03:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001246

INSTRUMENT REGISTERED TO: SOUTH DAYTONA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROMERO,RYAN,D
BREATH TEST OPERATOR AGENCY: South Daytona PD
ARRESTING OFFICER: TREAT,CLAY
ARRESTING OFFICER AGENCY: South Daytona PD
LAST AGENCY INSPECTION: 09/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 03:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: SEKUNNA

DATE/TIME OF TEST: 09/16/2019 05:03:36

Diagnostics Check	OK	04:56
Air Blank	0.000	04:57
Control Test	0.081	04:57
Air Blank	0.000	04:58
Subject Sample # 1	0.186	04:59
Breath Volume	2.722	04:59
Air Blank	0.000	04:59
Air Blank	0.000	05:01
Subject Sample # 2	0.188	05:02
Breath Volume	1.500	05:02
Air Blank	0.000	05:02
Control Test	0.082	05:03
Air Blank	0.000	05:03
Diagnostics Check	OK	05:03

Data Download Date/Time: 10/19/2019 02:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001246

INSTRUMENT REGISTERED TO: SOUTH DAYTONA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELODY, FRANCIS, D
BREATH TEST OPERATOR AGENCY: South Daytona PD
ARRESTING OFFICER: GRONDIN, BRYAN
ARRESTING OFFICER AGENCY: South Daytona PD
LAST AGENCY INSPECTION: 09/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: FACE

DATE/TIME OF TEST: 09/17/2019 00:57:54

Diagnostics Check	OK	00:51
Air Blank	0.000	00:52
Control Test	0.080	00:52
Air Blank	0.000	00:53
Subject Sample # 1	0.223	00:53
Breath Volume	2.542	00:53
Air Blank	0.000	00:54
Air Blank	0.000	00:55
Subject Sample # 2	0.221	00:56
Breath Volume	2.492	00:56
Air Blank	0.000	00:56
Control Test	0.081	00:57
Air Blank	0.000	00:57
Diagnostics Check	OK	00:57

Data Download Date/Time: 10/19/2019 02:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBIN,GASTON,A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: IRIZARRY,HEIDEN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 08/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 09/08/2019 01:07:35

Diagnostics Check	OK	01:01
Air Blank	0.000	01:01
Control Test	0.080	01:02
Air Blank	0.000	01:02
Subject Sample # 1	0.084	01:03
Breath Volume	2.156	01:03
Air Blank	0.000	01:03
Air Blank	0.000	01:05
Subject Sample # 2	0.081	01:06
Breath Volume	1.992	01:06
Air Blank	0.000	01:06
Control Test	0.081	01:07
Air Blank	0.000	01:07
Diagnostics Check	OK	01:07

Data Download Date/Time: 09/25/2019 21:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHAOUNI,K,
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: DAUGIRDAS,C
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 08/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: LOVERSO

DATE/TIME OF TEST: 09/13/2019 01:30:03

Diagnostics Check	OK	01:23
Air Blank	0.000	01:24
Control Test	0.080	01:24
Air Blank	0.000	01:25
Subject Sample # 1	0.000	01:25
Breath Volume	2.015	01:25
Air Blank	0.000	01:26
Air Blank	0.000	01:28
Subject Sample # 2	0.000	01:28
Breath Volume	1.375	01:28
Air Blank	0.000	01:29
Control Test	0.081	01:29
Air Blank	0.000	01:29
Diagnostics Check	OK	01:30

Data Download Date/Time: 09/25/2019 21:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBIN,GASTON,A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: ROBIN,GASTON
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 08/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 09/13/2019 22:42:53

Diagnostics Check	OK	22:36
Air Blank	0.000	22:37
Control Test	0.080	22:37
Air Blank	0.000	22:38
Subject Sample # 1	0.000	22:38
Breath Volume	2.804	22:38
Air Blank	0.000	22:39
Air Blank	0.000	22:41
Subject Sample # 2	0.000	22:41
Breath Volume	3.046	22:41
Air Blank	0.000	22:41
Control Test	0.081	22:42
Air Blank	0.000	22:42
Diagnostics Check	OK	22:42

Data Download Date/Time: 09/25/2019 21:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBIN,GASTON,A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: ROBIN,GASTON
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 08/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: SCHLONEGER

DATE/TIME OF TEST: 09/16/2019 23:58:57

Diagnostics Check	OK	23:50
Air Blank	0.000	23:50
Control Test	0.081	23:51
Air Blank	0.000	23:51
Subject Sample # 1	0.152	23:53
Breath Volume	2.730	23:53
Air Blank	0.000	23:54
Air Blank	0.000	23:56
Subject Sample # 2	0.143	23:57
Breath Volume	1.855	23:57
Air Blank	0.000	23:58
Control Test	0.081	23:58
Air Blank	0.000	23:58
Diagnostics Check	OK	23:58

Data Download Date/Time: 09/25/2019 21:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBIN,GASTON,A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: ROBIN,GASTON
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 08/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: CANAS

DATE/TIME OF TEST: 09/21/2019 00:39:07

Diagnostics Check	OK	00:31
Air Blank	0.000	00:31
Control Test	0.082	00:32
Air Blank	0.000	00:32
Subject Sample # 1	0.209	00:34
Breath Volume	1.492	00:34
Air Blank	0.000	00:34
Air Blank	0.000	00:36
Subject Sample # 2	0.206	00:37
Breath Volume	1.468	00:37
Air Blank	0.000	00:38
Control Test	0.081	00:38
Air Blank	0.000	00:39
Diagnostics Check	OK	00:39

Data Download Date/Time: 09/25/2019 21:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBIN,GASTON,A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: IRIZARRY,HEIDEN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 08/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: PINKOSKY

DATE/TIME OF TEST: 09/21/2019 03:31:37

Diagnostics Check	OK	03:24
Air Blank	0.000	03:25
Control Test	0.080	03:25
Air Blank	0.000	03:26
Subject Sample # 1	0.237	03:26
Breath Volume	2.324	03:26
Air Blank	0.000	03:27
Air Blank	0.000	03:29
Subject Sample # 2	0.234	03:30
Breath Volume	1.828	03:30
Air Blank	0.000	03:30
Control Test	0.080	03:31
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time: 09/25/2019 21:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBIN,GASTON,A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: GELL,RYAN
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 08/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: FLOWERS

DATE/TIME OF TEST: 09/21/2019 04:38:58

Diagnostics Check	OK	04:32
Air Blank	0.000	04:32
Control Test	0.081	04:32
Air Blank	0.000	04:33
Subject Sample # 1	0.143	04:34
Breath Volume	1.855	04:34
Air Blank	0.000	04:35
Air Blank	0.000	04:36
Subject Sample # 2	0.146	04:37
Breath Volume	2.210	04:37
Air Blank	0.000	04:38
Control Test	0.081	04:38
Air Blank	0.000	04:38
Diagnostics Check	OK	04:38

Data Download Date/Time: 09/25/2019 21:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BRYAN,ADAM
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 08/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: GEORGES

DATE/TIME OF TEST: 09/01/2019 05:52:50

Diagnostics Check	OK	05:43
Air Blank	0.000	05:43
Control Test	0.080	05:44
Air Blank	0.000	05:44
Subject Sample # 1	0.181	05:47
Breath Volume	1.625	05:47
Air Blank	0.000	05:48
Air Blank	0.000	05:49
Subject Sample # 2	0.179	05:51
Breath Volume	1.484	05:51
Air Blank	0.000	05:51
Control Test	0.077	05:52
Air Blank	0.000	05:52
Diagnostics Check	OK	05:52

Data Download Date/Time: 09/04/2019 14:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HARRIS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 08/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:24

SUBJECT'S LAST NAME: OLSON

DATE/TIME OF TEST: 09/01/2019 17:56:44

Diagnostics Check	OK	17:47
Air Blank	0.000	17:47
Control Test	0.080	17:48
Air Blank	0.000	17:48
Subject Sample # 1	0.216	17:49
Breath Volume	4.789	17:49
Air Blank	0.000	17:49
Air Blank	0.000	17:51
Subject Sample # 2	0.194	17:52
Breath Volume	1.789	17:52
Air Blank	0.000	17:52
Air Blank	0.000	17:54
Subject Sample # 3	0.215	17:55
Breath Volume	4.523	17:55
Air Blank	0.000	17:55
Control Test	0.078	17:56
Air Blank	0.000	17:56
Diagnostics Check	OK	17:56

Data Download Date/Time: 09/04/2019 14:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS, ARMAND, G
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HARWOOD, ARMANDO
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 08/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:32

SUBJECT'S LAST NAME: MANZANAREZ

DATE/TIME OF TEST: 09/03/2019 18:02:38

Diagnostics Check	OK	17:56
Air Blank	0.000	17:56
Control Test	0.080	17:57
Air Blank	0.000	17:57
Subject Sample # 1	0.169	17:58
Breath Volume	2.800	17:58
Air Blank	0.000	17:58
Air Blank	0.000	18:00
Subject Sample # 2	0.167	18:01
Breath Volume	2.808	18:01
Air Blank	0.000	18:01
Control Test	0.079	18:02
Air Blank	0.000	18:02
Diagnostics Check	OK	18:02

Data Download Date/Time: 09/04/2019 14:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CORSO,ANTHONY
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 08/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 09/06/2019 03:45:55

Diagnostics Check	OK	03:38
Air Blank	0.000	03:39
Control Test	0.080	03:39
Air Blank	0.000	03:39
Subject Sample # 1	0.242	03:41
Breath Volume	1.273	03:41
Air Blank	0.000	03:42
Air Blank	0.000	03:43
Subject Sample # 2	0.250	03:44
Breath Volume	1.406	03:44
Air Blank	0.000	03:45
Control Test	0.078	03:45
Air Blank	0.000	03:45
Diagnostics Check	OK	03:45

Data Download Date/Time: 09/11/2019 11:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MICHAEL, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AGOSTO, ALBA
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 07:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:45

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 09/08/2019 08:20:20

Diagnostics Check	OK	08:13
Air Blank	0.000	08:14
Control Test	0.080	08:14
Air Blank	0.000	08:15
Subject Sample # 1	0.181	08:15
Breath Volume	1.164	08:15
Air Blank	0.000	08:16
Air Blank	0.000	08:18
Subject Sample # 2	0.180	08:18
Breath Volume	2.214	08:18
Air Blank	0.000	08:19
Control Test	0.079	08:19
Air Blank	0.000	08:20
Diagnostics Check	OK	08:20

Data Download Date/Time: 09/11/2019 11:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELVILLE,KELLY,M
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DOMINGUEZ,JAYMES
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 08/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 08:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:36

SUBJECT'S LAST NAME: VIEIRA

DATE/TIME OF TEST: 09/10/2019 09:05:52

Diagnostics Check	OK	09:00
Air Blank	0.000	09:00
Control Test	0.080	09:00
Air Blank	0.000	09:01
Subject Sample # 1	0.105	09:01
Breath Volume	1.480	09:01
Air Blank	0.000	09:02
Air Blank	0.000	09:04
Subject Sample # 2	0.104	09:04
Breath Volume	1.835	09:04
Air Blank	0.000	09:04
Control Test	0.079	09:05
Air Blank	0.000	09:05
Diagnostics Check	OK	09:05

Data Download Date/Time: 09/11/2019 11:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RODRIGUEZ, JORGE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: RAMPERSAD

DATE/TIME OF TEST: 09/11/2019 18:42:10

Diagnostics Check	OK	18:35
Air Blank	0.000	18:36
Control Test	0.079	18:36
Air Blank	0.000	18:37
Subject Sample # 1	0.139	18:37
Breath Volume	1.933	18:37
Air Blank	0.000	18:38
Air Blank	0.000	18:40
Subject Sample # 2	0.152	18:40
Breath Volume	1.875	18:40
Air Blank	0.000	18:41
Control Test	0.079	18:41
Air Blank	0.000	18:42
Diagnostics Check	OK	18:42

Data Download Date/Time: 09/19/2019 16:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SMOOTZ,EMBER
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: ZUKOWSKI

DATE/TIME OF TEST: 09/11/2019 21:27:01

Diagnostics Check	OK	21:20
Air Blank	0.000	21:20
Control Test	0.079	21:20
Air Blank	0.000	21:21
Subject Sample # 1	0.170	21:22
Breath Volume	1.941	21:22
Air Blank	0.000	21:23
Air Blank	0.000	21:25
Subject Sample # 2	0.165	21:25
Breath Volume	1.710	21:25
Air Blank	0.000	21:26
Control Test	0.077	21:26
Air Blank	0.000	21:26
Diagnostics Check	OK	21:27

Data Download Date/Time: 09/19/2019 16:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HARWOOD,ARMANDO
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: ALGARIN

DATE/TIME OF TEST: 09/12/2019 01:23:30

Diagnostics Check	OK	01:16
Air Blank	0.000	01:17
Control Test	0.080	01:17
Air Blank	0.000	01:17
Subject Sample # 1	0.146	01:18
Breath Volume	1.406	01:18
Air Blank	0.000	01:19
Air Blank	0.000	01:21
Subject Sample # 2	0.143	01:21
Breath Volume	2.335	01:21
Air Blank	0.000	01:22
Control Test	0.079	01:22
Air Blank	0.000	01:23
Diagnostics Check	OK	01:23

Data Download Date/Time: 09/19/2019 16:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS, ARMAND, G
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ADAMS, JAMES
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:37

SUBJECT'S LAST NAME: STOCKER

DATE/TIME OF TEST: 09/12/2019 17:15:16

Diagnostics Check	OK	17:09
Air Blank	0.000	17:09
Control Test	0.080	17:09
Air Blank	0.000	17:10
Subject Sample # 1	0.000	17:10
Breath Volume	2.839	17:10
Air Blank	0.000	17:11
Air Blank	0.000	17:13
Subject Sample # 2	0.000	17:13
Breath Volume	2.261	17:13
Air Blank	0.000	17:14
Control Test	0.080	17:14
Air Blank	0.000	17:15
Diagnostics Check	OK	17:15

Data Download Date/Time: 09/19/2019 16:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WISE,IAN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:28

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 09/13/2019 21:56:17

Diagnostics Check	OK	21:50
Air Blank	0.000	21:50
Control Test	0.080	21:51
Air Blank	0.000	21:51
Subject Sample # 1	0.000	21:52
Breath Volume	1.515	21:52
Air Blank	0.000	21:52
Air Blank	0.000	21:54
Subject Sample # 2	0.000	21:54
Breath Volume	2.769	21:54
Air Blank	0.000	21:55
Control Test	0.080	21:55
Air Blank	0.000	21:56
Diagnostics Check	OK	21:56

Data Download Date/Time: 09/19/2019 16:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MORALES,JOSE
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: SANTUCCI

DATE/TIME OF TEST: 09/14/2019 00:30:06

Diagnostics Check	OK	00:23
Air Blank	0.000	00:23
Control Test	0.080	00:24
Air Blank	0.000	00:24
Subject Sample # 1	0.000	00:25
Breath Volume	2.175	00:25
Air Blank	0.000	00:26
Air Blank	0.000	00:28
Subject Sample # 2	0.000	00:28
Breath Volume	1.777	00:28
Air Blank	0.000	00:29
Control Test	0.080	00:29
Air Blank	0.000	00:30
Diagnostics Check	OK	00:30

Data Download Date/Time: 09/19/2019 16:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GUY,STEVEN
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: MCDANIEL

DATE/TIME OF TEST: 09/14/2019 01:28:39

Diagnostics Check	OK	01:19
Air Blank	0.000	01:20
Control Test	0.080	01:20
Air Blank	0.000	01:20
Subject Sample # 1	0.199	01:23
Breath Volume	2.035	01:23
Air Blank	0.000	01:23
Air Blank	0.000	01:25
Subject Sample # 2	0.195	01:27
Breath Volume	1.605	01:27
Air Blank	0.000	01:27
Control Test	0.079	01:28
Air Blank	0.000	01:28
Diagnostics Check	OK	01:28

Data Download Date/Time: 09/19/2019 16:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VICKERS,TRACY
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: ESPADA SANCHEZ

DATE/TIME OF TEST: 09/14/2019 04:13:52

Diagnostics Check	OK	04:07
Air Blank	0.000	04:08
Control Test	0.080	04:08
Air Blank	0.000	04:08
Subject Sample # 1	0.057	04:09
Breath Volume	2.625	04:09
Air Blank	0.000	04:10
Air Blank	0.000	04:11
Subject Sample # 2	0.055	04:12
Breath Volume	2.640	04:12
Air Blank	0.000	04:12
Control Test	0.080	04:13
Air Blank	0.000	04:13
Diagnostics Check	OK	04:13

Data Download Date/Time: 09/19/2019 16:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: REDWAN,ALFREDO
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:25

SUBJECT'S LAST NAME: CABALLERO CABALLERO

DATE/TIME OF TEST: 09/14/2019 05:56:53

Diagnostics Check	OK	05:50
Air Blank	0.000	05:50
Control Test	0.081	05:51
Air Blank	0.000	05:51
Subject Sample # 1	0.085	05:52
Breath Volume	1.101	05:52
Air Blank	0.000	05:52
Air Blank	0.000	05:54
Subject Sample # 2	0.084	05:55
Breath Volume	1.632	05:55
Air Blank	0.000	05:55
Control Test	0.079	05:56
Air Blank	0.000	05:56
Diagnostics Check	OK	05:56

Data Download Date/Time: 09/19/2019 16:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MICHAEL, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WELSCHINGER, MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:10

SUBJECT'S LAST NAME: DELAUTER

DATE/TIME OF TEST: 09/14/2019 17:39:24

Diagnostics Check	OK	17:33
Air Blank	0.000	17:33
Control Test	0.080	17:33
Air Blank	0.000	17:34
Subject Sample # 1	0.206	17:34
Breath Volume	3.703	17:34
Air Blank	0.000	17:35
Air Blank	0.000	17:37
Subject Sample # 2	0.211	17:37
Breath Volume	4.281	17:37
Air Blank	0.000	17:38
Control Test	0.079	17:38
Air Blank	0.000	17:39
Diagnostics Check	OK	17:39

Data Download Date/Time: 09/19/2019 16:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PAULINO,JOSE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: VALENCIA

DATE/TIME OF TEST: 09/14/2019 21:53:04

Diagnostics Check	OK	21:46
Air Blank	0.000	21:46
Control Test	0.080	21:47
Air Blank	0.000	21:47
Subject Sample # 1	0.121	21:48
Breath Volume	2.109	21:48
Air Blank	0.000	21:49
Air Blank	0.000	21:51
Subject Sample # 2	0.125	21:51
Breath Volume	1.917	21:51
Air Blank	0.000	21:52
Control Test	0.078	21:52
Air Blank	0.000	21:52
Diagnostics Check	OK	21:53

Data Download Date/Time: 09/19/2019 16:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WISE,IAN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: MEDINA BONILLA

DATE/TIME OF TEST: 09/15/2019 03:44:39

Diagnostics Check	OK	03:31	
Air Blank	0.000	03:31	
Control Test	0.079	03:31	
Air Blank	0.000	03:32	
Subject Sample	# 1 0.101	03:35	Volume Not Met
Breath Volume	0.128	03:35	
Air Blank	0.000	03:36	
Air Blank	0.000	03:37	
Subject Sample	# 2 0.099	03:39	
Breath Volume	1.574	03:39	
Air Blank	0.000	03:39	
Air Blank	0.000	03:41	
Subject Sample	# 3 0.099	03:43	
Breath Volume	1.109	03:43	
Air Blank	0.000	03:43	
Control Test	0.078	03:44	
Air Blank	0.000	03:44	
Diagnostics Check	OK	03:44	

Data Download Date/Time: 09/19/2019 16:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ERICKSON,JORDAN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: CLASS LOPEZ

DATE/TIME OF TEST: 09/15/2019 04:49:19

Diagnostics Check	OK	04:42
Air Blank	0.000	04:42
Control Test	0.079	04:43
Air Blank	0.000	04:43
Subject Sample # 1	0.088	04:44
Breath Volume	1.636	04:44
Air Blank	0.000	04:45
Air Blank	0.000	04:46
Subject Sample # 2	0.083	04:47
Breath Volume	1.480	04:47
Air Blank	0.000	04:48
Control Test	0.079	04:48
Air Blank	0.000	04:49
Diagnostics Check	OK	04:49

Data Download Date/Time: 09/19/2019 16:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GUEVARA,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:46

SUBJECT'S LAST NAME: PORTWOOD VELEZ

DATE/TIME OF TEST: 09/15/2019 05:16:02

Diagnostics Check	OK	05:09
Air Blank	0.000	05:10
Control Test	0.078	05:10
Air Blank	0.000	05:11
Subject Sample # 1	0.165	05:11
Breath Volume	1.285	05:11
Air Blank	0.000	05:12
Air Blank	0.000	05:13
Subject Sample # 2	0.164	05:14
Breath Volume	2.535	05:14
Air Blank	0.000	05:15
Control Test	0.079	05:15
Air Blank	0.000	05:15
Diagnostics Check	OK	05:16

Data Download Date/Time: 09/19/2019 16:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELVILLE,KELLY,M
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CADY,PETER
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:55

SUBJECT'S LAST NAME: MORSE

DATE/TIME OF TEST: 09/17/2019 17:31:53

Diagnostics Check	OK	17:19	
Air Blank	0.000	17:20	
Control Test	0.080	17:20	
Air Blank	0.000	17:21	
Subject Sample	# 1 0.156	17:24	Volume Not Met
Breath Volume	0.175	17:24	
Air Blank	0.000	17:25	
Air Blank	0.000	17:26	
Subject Sample	# 2 0.166	17:27	
Breath Volume	1.859	17:27	
Air Blank	0.000	17:27	
Air Blank	0.000	17:29	
Subject Sample	# 3 0.150	17:30	
Breath Volume	1.570	17:30	
Air Blank	0.000	17:30	
Control Test	0.077	17:31	
Air Blank	0.000	17:31	
Diagnostics Check	OK	17:31	

Data Download Date/Time: 09/19/2019 16:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HARRIS,BRIAN
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:14

SUBJECT'S LAST NAME: VELASQUEZPEREZ

DATE/TIME OF TEST: 09/17/2019 20:45:17

Diagnostics Check	OK	20:39
Air Blank	0.000	20:39
Control Test	0.079	20:39
Air Blank	0.000	20:40
Subject Sample # 1	0.174	20:40
Breath Volume	2.195	20:40
Air Blank	0.000	20:41
Air Blank	0.000	20:43
Subject Sample # 2	0.169	20:43
Breath Volume	2.156	20:43
Air Blank	0.000	20:44
Control Test	0.079	20:44
Air Blank	0.000	20:45
Diagnostics Check	OK	20:45

Data Download Date/Time: 09/19/2019 16:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JONES,EUGENE
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: CARPENTER

DATE/TIME OF TEST: 09/18/2019 02:20:30

Diagnostics Check	OK	02:11
Air Blank	0.000	02:11
Control Test	0.080	02:11
Air Blank	0.000	02:12
Subject Sample # 1	0.279	02:12
Breath Volume	3.535	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample # 2	0.239	02:15
Breath Volume	1.511	02:15
Air Blank	0.000	02:16
Air Blank	0.000	02:18
Subject Sample # 3	0.264	02:18
Breath Volume	3.238	02:18
Air Blank	0.000	02:19
Control Test	0.079	02:19
Air Blank	0.000	02:20
Diagnostics Check	OK	02:20

Data Download Date/Time: 09/19/2019 16:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HALLOCK,STEPHEN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: CASIMER

DATE/TIME OF TEST: 09/19/2019 00:58:17

Diagnostics Check	OK	00:51
Air Blank	0.000	00:52
Control Test	0.080	00:52
Air Blank	0.000	00:53
Subject Sample # 1	0.350	00:53
Breath Volume	4.359	00:53
Air Blank	0.000	00:54
Air Blank	0.000	00:56
Subject Sample # 2	0.359	00:56
Breath Volume	4.335	00:56
Air Blank	0.000	00:57
Control Test	0.078	00:57
Air Blank	0.000	00:58
Diagnostics Check	OK	00:58

Data Download Date/Time: 09/19/2019 16:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: REMBISLICARI,JAZMIN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: ELLIS

DATE/TIME OF TEST: 09/21/2019 03:33:42

Diagnostics Check	OK	03:23
Air Blank	0.000	03:24
Control Test	0.079	03:24
Air Blank	0.000	03:24
Subject Sample # 1	0.103	03:27
Breath Volume	1.285	03:27
Air Blank	0.000	03:28
Air Blank	0.000	03:30
Subject Sample # 2	0.094	03:32
Breath Volume	1.207	03:32
Air Blank	0.000	03:32
Control Test	0.077	03:33
Air Blank	0.000	03:33
Diagnostics Check	OK	03:33

Data Download Date/Time: 09/26/2019 14:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LOPEZ,OSCAR
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:41

SUBJECT'S LAST NAME: KELLEY

DATE/TIME OF TEST: 09/22/2019 21:11:01

Diagnostics Check	OK	21:04
Air Blank	0.000	21:04
Control Test	0.079	21:05
Air Blank	0.000	21:05
Subject Sample # 1	0.192	21:06
Breath Volume	1.320	21:06
Air Blank	0.000	21:07
Air Blank	0.000	21:08
Subject Sample # 2	0.212	21:09
Breath Volume	1.351	21:09
Air Blank	0.000	21:10
Control Test	0.078	21:10
Air Blank	0.000	21:10
Diagnostics Check	OK	21:11

Data Download Date/Time: 09/26/2019 14:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELVILLE,KELLY,M
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MARTINEZ,YAISEL
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 07:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:18

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 09/23/2019 08:50:49

Diagnostics Check	OK	08:44
Air Blank	0.000	08:44
Control Test	0.079	08:44
Air Blank	0.000	08:45
Subject Sample # 1	0.000	08:46
Breath Volume	1.546	08:46
Air Blank	0.000	08:46
Air Blank	0.000	08:48
Subject Sample # 2	0.000	08:49
Breath Volume	1.695	08:49
Air Blank	0.000	08:49
Control Test	0.080	08:50
Air Blank	0.000	08:50
Diagnostics Check	OK	08:50

Data Download Date/Time: 09/26/2019 14:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BRAZEE,COLE
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 09/24/2019 01:35:38

Diagnostics Check	OK	01:29
Air Blank	0.000	01:29
Control Test	0.080	01:30
Air Blank	0.000	01:30
Subject Sample # 1	0.194	01:31
Breath Volume	2.179	01:31
Air Blank	0.000	01:31
Air Blank	0.000	01:33
Subject Sample # 2	0.198	01:34
Breath Volume	2.410	01:34
Air Blank	0.000	01:34
Control Test	0.079	01:35
Air Blank	0.000	01:35
Diagnostics Check	OK	01:35

Data Download Date/Time: 09/26/2019 14:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HOPKINS,JEFFREY
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: BRENNER

DATE/TIME OF TEST: 09/24/2019 03:20:39

Diagnostics Check	OK	03:14
Air Blank	0.000	03:14
Control Test	0.080	03:15
Air Blank	0.000	03:15
Subject Sample # 1	0.217	03:16
Breath Volume	2.425	03:16
Air Blank	0.000	03:16
Air Blank	0.000	03:18
Subject Sample # 2	0.215	03:19
Breath Volume	2.191	03:19
Air Blank	0.000	03:19
Control Test	0.079	03:20
Air Blank	0.000	03:20
Diagnostics Check	OK	03:20

Data Download Date/Time: 09/26/2019 14:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AMERICA, DEVA
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: VAUGHN

DATE/TIME OF TEST: 09/24/2019 23:52:15

Diagnostics Check	OK	23:45
Air Blank	0.000	23:46
Control Test	0.079	23:46
Air Blank	0.000	23:47
Subject Sample # 1	0.178	23:47
Breath Volume	1.261	23:47
Air Blank	0.000	23:48
Air Blank	0.000	23:50
Subject Sample # 2	0.171	23:50
Breath Volume	1.406	23:50
Air Blank	0.000	23:51
Control Test	0.079	23:51
Air Blank	0.000	23:52
Diagnostics Check	OK	23:52

Data Download Date/Time: 09/26/2019 14:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PEERS,MICHAEL
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: COOKE

DATE/TIME OF TEST: 09/25/2019 00:23:14

Diagnostics Check	OK	00:16
Air Blank	0.000	00:16
Control Test	0.080	00:17
Air Blank	0.000	00:17
Subject Sample # 1	0.209	00:18
Breath Volume	1.394	00:18
Air Blank	0.000	00:19
Air Blank	0.000	00:21
Subject Sample # 2	0.210	00:21
Breath Volume	1.617	00:21
Air Blank	0.000	00:22
Control Test	0.077	00:22
Air Blank	0.000	00:23
Diagnostics Check	OK	00:23

Data Download Date/Time: 09/26/2019 14:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WISE,IAN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: SALGUEROREYES

DATE/TIME OF TEST: 09/25/2019 01:58:09

Diagnostics Check	OK	01:51
Air Blank	0.000	01:51
Control Test	0.079	01:51
Air Blank	0.000	01:52
Subject Sample # 1	0.144	01:53
Breath Volume	2.312	01:53
Air Blank	0.000	01:54
Air Blank	0.000	01:55
Subject Sample # 2	0.140	01:56
Breath Volume	3.277	01:56
Air Blank	0.000	01:57
Control Test	0.079	01:57
Air Blank	0.000	01:58
Diagnostics Check	OK	01:58

Data Download Date/Time: 09/26/2019 14:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MARTINEZ,ERNESTO
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 09/25/2019 03:53:54

Diagnostics Check	OK	03:40	
Air Blank	0.000	03:41	
Control Test	0.079	03:41	
Air Blank	0.000	03:42	
Subject Sample	# 1 0.181	03:43	
Breath Volume	1.382	03:43	
Air Blank	0.000	03:44	
Air Blank	0.000	03:46	
Subject Sample	# 2 0.147	03:49	Volume Not Met
Breath Volume	0.847	03:49	
Air Blank	0.000	03:50	
Air Blank	0.000	03:51	
Subject Sample	# 3 0.195	03:52	
Breath Volume	2.769	03:52	
Air Blank	0.000	03:52	
Control Test	0.077	03:53	
Air Blank	0.000	03:53	
Diagnostics Check	OK	03:53	

Data Download Date/Time: 09/26/2019 14:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RAMEY,MIKE
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 09/25/2019 04:11:27

Diagnostics Check	OK	04:04
Air Blank	0.000	04:04
Control Test	0.079	04:05
Air Blank	0.000	04:05
Subject Sample # 1	0.204	04:06
Breath Volume	1.296	04:06
Air Blank	0.000	04:06
Air Blank	0.000	04:08
Subject Sample # 2	0.215	04:09
Breath Volume	1.632	04:09
Air Blank	0.000	04:10
Control Test	0.079	04:10
Air Blank	0.000	04:11
Diagnostics Check	OK	04:11

Data Download Date/Time: 09/26/2019 14:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELVILLE,KELLY,M
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ELLSWORTH,PETER
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:28

SUBJECT'S LAST NAME: BATHURST

DATE/TIME OF TEST: 09/25/2019 16:59:06

Diagnostics Check	OK	16:52
Air Blank	0.000	16:53
Control Test	0.079	16:53
Air Blank	0.000	16:54
Subject Sample # 1	0.181	16:54
Breath Volume	2.816	16:54
Air Blank	0.000	16:55
Air Blank	0.000	16:57
Subject Sample # 2	0.176	16:57
Breath Volume	2.703	16:57
Air Blank	0.000	16:58
Control Test	0.078	16:58
Air Blank	0.000	16:59
Diagnostics Check	OK	16:59

Data Download Date/Time: 09/26/2019 14:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MENDEZ,JOHN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: MONJEROMAN

DATE/TIME OF TEST: 09/25/2019 23:47:41

Diagnostics Check	OK	23:40
Air Blank	0.000	23:41
Control Test	0.080	23:41
Air Blank	0.000	23:42
Subject Sample # 1	0.185	23:42
Breath Volume	1.328	23:42
Air Blank	0.000	23:43
Air Blank	0.000	23:45
Subject Sample # 2	0.188	23:46
Breath Volume	3.105	23:46
Air Blank	0.000	23:46
Control Test	0.079	23:47
Air Blank	0.000	23:47
Diagnostics Check	OK	23:47

Data Download Date/Time: 09/26/2019 14:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MARTINEZ,ERNESTO
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 09/26/2019 03:26:08

Diagnostics Check	OK	03:19
Air Blank	0.000	03:20
Control Test	0.079	03:20
Air Blank	0.000	03:21
Subject Sample # 1	0.105	03:21
Breath Volume	1.648	03:21
Air Blank	0.000	03:22
Air Blank	0.000	03:24
Subject Sample # 2	0.103	03:24
Breath Volume	1.949	03:24
Air Blank	0.000	03:25
Control Test	0.079	03:25
Air Blank	0.000	03:26
Diagnostics Check	OK	03:26

Data Download Date/Time: 09/26/2019 14:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RODRIGUEZ,JORGE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: UMSCHIED

DATE/TIME OF TEST: 09/27/2019 03:27:34

Diagnostics Check	OK	03:21
Air Blank	0.000	03:21
Control Test	0.079	03:22
Air Blank	0.000	03:22
Subject Sample # 1	0.171	03:23
Breath Volume	2.460	03:23
Air Blank	0.000	03:23
Air Blank	0.000	03:25
Subject Sample # 2	0.167	03:26
Breath Volume	2.519	03:26
Air Blank	0.000	03:26
Control Test	0.079	03:26
Air Blank	0.000	03:27
Diagnostics Check	OK	03:27

Data Download Date/Time: 10/03/2019 17:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SIMPSON,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: STJOHN

DATE/TIME OF TEST: 09/28/2019 01:52:37

Diagnostics Check	OK	01:45
Air Blank	0.000	01:46
Control Test	0.080	01:46
Air Blank	0.000	01:46
Subject Sample # 1	0.141	01:48
Breath Volume	3.578	01:48
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample # 2	0.144	01:51
Breath Volume	3.585	01:51
Air Blank	0.000	01:51
Control Test	0.079	01:52
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

Data Download Date/Time: 10/03/2019 17:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GUEVARA,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: EAGAN

DATE/TIME OF TEST: 09/28/2019 02:45:19

Diagnostics Check	OK	02:38
Air Blank	0.000	02:39
Control Test	0.080	02:39
Air Blank	0.000	02:40
Subject Sample # 1	0.131	02:40
Breath Volume	1.632	02:40
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.136	02:43
Breath Volume	2.500	02:43
Air Blank	0.000	02:44
Control Test	0.080	02:44
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 10/03/2019 17:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JOHNS,MARK
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 09/28/2019 04:57:57

Diagnostics Check	OK	04:46	
Air Blank	0.000	04:47	
Control Test	0.080	04:47	
Air Blank	0.000	04:47	
Subject Sample	# 1	04:50	No Sample Provided
Breath Volume	0.000	04:50	
Air Blank	0.000	04:51	
Air Blank	0.000	04:53	
Subject Sample	# 2	04:56	No Sample Provided
Breath Volume	0.000	04:56	
Air Blank	0.000	04:57	
Control Test	0.080	04:57	
Air Blank	0.000	04:57	
Diagnostics Check	OK	04:57	

Data Download Date/Time: 10/03/2019 17:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PRICE,FORREST
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:35

SUBJECT'S LAST NAME: ISAAC

DATE/TIME OF TEST: 09/28/2019 17:06:28

Diagnostics Check	OK	16:59
Air Blank	0.000	17:00
Control Test	0.080	17:00
Air Blank	0.000	17:01
Subject Sample # 1	0.203	17:01
Breath Volume	1.585	17:01
Air Blank	0.000	17:02
Air Blank	0.000	17:04
Subject Sample # 2	0.214	17:04
Breath Volume	1.410	17:04
Air Blank	0.000	17:05
Control Test	0.078	17:05
Air Blank	0.000	17:06
Diagnostics Check	OK	17:06

Data Download Date/Time: 10/03/2019 17:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HOPKINS,JEFFREY
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: NGUYEN

DATE/TIME OF TEST: 09/28/2019 21:21:11

Diagnostics Check	OK	21:15
Air Blank	0.000	21:15
Control Test	0.079	21:15
Air Blank	0.000	21:16
Subject Sample # 1	0.164	21:16
Breath Volume	3.082	21:16
Air Blank	0.000	21:17
Air Blank	0.000	21:19
Subject Sample # 2	0.160	21:19
Breath Volume	2.378	21:19
Air Blank	0.000	21:20
Control Test	0.078	21:20
Air Blank	0.000	21:21
Diagnostics Check	OK	21:21

Data Download Date/Time: 10/03/2019 17:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PENDER,KHAYREE
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: VILLARREALARCILA

DATE/TIME OF TEST: 09/28/2019 21:51:46

Diagnostics Check	OK	21:45
Air Blank	0.000	21:45
Control Test	0.080	21:46
Air Blank	0.000	21:46
Subject Sample # 1	0.162	21:47
Breath Volume	2.386	21:47
Air Blank	0.000	21:47
Air Blank	0.000	21:49
Subject Sample # 2	0.162	21:50
Breath Volume	2.277	21:50
Air Blank	0.000	21:50
Control Test	0.078	21:51
Air Blank	0.000	21:51
Diagnostics Check	OK	21:51

Data Download Date/Time: 10/03/2019 17:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WELSCHINGER,MICKAEL
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: SCHEIN

DATE/TIME OF TEST: 09/28/2019 22:50:43

Diagnostics Check	OK	22:43
Air Blank	0.000	22:44
Control Test	0.079	22:44
Air Blank	0.000	22:44
Subject Sample # 1	0.222	22:46
Breath Volume	1.277	22:46
Air Blank	0.000	22:47
Air Blank	0.000	22:48
Subject Sample # 2	0.202	22:49
Breath Volume	1.523	22:49
Air Blank	0.000	22:49
Control Test	0.077	22:50
Air Blank	0.000	22:50
Diagnostics Check	OK	22:50

Data Download Date/Time: 10/03/2019 17:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MILLER,JACOB
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: GOMEZ BRAVO

DATE/TIME OF TEST: 09/29/2019 21:06:04

Diagnostics Check	OK	20:59
Air Blank	0.000	21:00
Control Test	0.079	21:00
Air Blank	0.000	21:01
Subject Sample # 1	0.312	21:01
Breath Volume	2.414	21:01
Air Blank	0.000	21:02
Air Blank	0.000	21:04
Subject Sample # 2	0.303	21:04
Breath Volume	2.226	21:04
Air Blank	0.000	21:05
Control Test	0.079	21:05
Air Blank	0.000	21:05
Diagnostics Check	OK	21:06

Data Download Date/Time: 10/03/2019 17:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RAMIREZ,MATTHEW
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: LOPEZGARCIA

DATE/TIME OF TEST: 09/30/2019 00:01:53

Diagnostics Check	OK	23:54
Air Blank	0.000	23:54
Control Test	0.079	23:54
Air Blank	0.000	23:55
Subject Sample # 2	0.246	23:57
Breath Volume	1.960	23:57
Air Blank	0.000	23:57
Air Blank	0.000	23:59
Subject Sample # 1	0.254	00:00
Breath Volume	2.636	00:00
Air Blank	0.000	00:00
Control Test	0.077	00:01
Air Blank	0.000	00:01
Diagnostics Check	OK	00:01

Data Download Date/Time: 10/03/2019 17:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

INSTRUMENT REGISTERED TO: COCOA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,JAMES,H
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: ZUNIGA,C
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 08/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: CARROLL

DATE/TIME OF TEST: 09/12/2019 23:57:45

Diagnostics Check	OK	23:54	
Air Blank	0.000	23:55	
Control Test	0.079	23:55	
Air Blank	0.000	23:56	
Subject Sample	# 1	23:56	Subject Test Refused
Breath Volume	0.000	23:56	
Air Blank	0.000	23:56	
Control Test	0.079	23:57	
Air Blank	0.000	23:57	
Diagnostics Check	OK	23:57	

Data Download Date/Time: 09/24/2019 23:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

INSTRUMENT REGISTERED TO: COCOA P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PICHECO,JOHN,A

BREATH TEST OPERATOR AGENCY: Cocoa PD

ARRESTING OFFICER: BAUMANN,MATTHEW

ARRESTING OFFICER AGENCY: Rockledge PD

LAST AGENCY INSPECTION: 08/28/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 02618080A3

DRY GAS EXPIRY DATE: 03/05/2020

VIOLATION CODE: DUI

ARREST TIME: 02:37

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: MCLAUGHLIN

DATE/TIME OF TEST: 09/17/2019 03:43:37

Diagnostics Check	OK	03:37
Air Blank	0.000	03:37
Control Test	0.080	03:37
Air Blank	0.000	03:38
Subject Sample	# 1 0.190	03:38
Breath Volume	2.667	03:38
Air Blank	0.000	03:39
Air Blank	0.000	03:41
Subject Sample	# 2 0.186	03:42
Breath Volume	2.613	03:42
Air Blank	0.000	03:42
Control Test	0.079	03:43
Air Blank	0.000	03:43
Diagnostics Check	OK	03:43

Data Download Date/Time: 09/24/2019 23:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,ANTHONY,M
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: POND,M
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 08/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 09/01/2019 03:21:46

Diagnostics Check	OK	03:15
Air Blank	0.000	03:15
Control Test	0.080	03:15
Air Blank	0.000	03:16
Subject Sample # 1	0.130	03:16
Breath Volume	1.441	03:16
Air Blank	0.000	03:17
Air Blank	0.000	03:19
Subject Sample # 2	0.121	03:20
Breath Volume	1.234	03:20
Air Blank	0.000	03:20
Control Test	0.080	03:21
Air Blank	0.000	03:21
Diagnostics Check	OK	03:21

Data Download Date/Time: 09/30/2019 13:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,ANTHONY,M
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: VURRO,M
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 08/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 09/01/2019 03:57:28

Diagnostics Check	OK	03:48
Air Blank	0.000	03:48
Control Test	0.080	03:49
Air Blank	0.000	03:49
Subject Sample # 1	0.091	03:52
Breath Volume	1.468	03:52
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample # 2	0.084	03:55
Breath Volume	1.906	03:55
Air Blank	0.000	03:56
Control Test	0.078	03:56
Air Blank	0.000	03:57
Diagnostics Check	OK	03:57

Data Download Date/Time: 09/30/2019 13:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILBIN,SKYLER,J
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: DUMMER,LUKE
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 08/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: HOFFMAN

DATE/TIME OF TEST: 09/11/2019 00:43:55

Diagnostics Check	OK	00:30	
Air Blank	0.000	00:31	
Control Test	0.080	00:31	
Air Blank	0.000	00:32	
Subject Sample	# 1 0.112	00:35	Volume Not Met
Breath Volume	0.042	00:35	
Air Blank	0.000	00:36	
Air Blank	0.000	00:37	
Subject Sample	# 2 0.123	00:38	
Breath Volume	1.437	00:38	
Air Blank	0.000	00:39	
Air Blank	0.000	00:41	
Subject Sample	# 3 0.120	00:42	
Breath Volume	1.359	00:42	
Air Blank	0.000	00:43	
Control Test	0.078	00:43	
Air Blank	0.000	00:43	
Diagnostics Check	OK	00:43	

Data Download Date/Time: 09/30/2019 13:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,B,D
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: PEREZ,V
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 08/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 15:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:21

SUBJECT'S LAST NAME: BRADY

DATE/TIME OF TEST: 09/13/2019 16:45:36

Diagnostics Check	OK	16:42	
Air Blank	0.000	16:42	
Control Test	0.080	16:43	
Air Blank	0.000	16:43	
Subject Sample	# 1	16:44	Slope Not Met
Breath Volume	0.847	16:44	
Air Blank	0.000	16:44	
Control Test	0.080	16:45	
Air Blank	0.000	16:45	
Diagnostics Check	OK	16:45	

Data Download Date/Time: 09/30/2019 13:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,B,D
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: PEREZ,V
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 08/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 15:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:21

SUBJECT'S LAST NAME: BRADY

DATE/TIME OF TEST: 09/13/2019 16:53:55

Diagnostics Check	OK	16:47
Air Blank	0.000	16:47
Control Test	0.080	16:48
Air Blank	0.000	16:48
Subject Sample # 1	0.240	16:49
Breath Volume	1.933	16:49
Air Blank	0.000	16:49
Air Blank	0.000	16:51
Subject Sample # 2	0.238	16:52
Breath Volume	1.609	16:52
Air Blank	0.000	16:53
Control Test	0.079	16:53
Air Blank	0.000	16:53
Diagnostics Check	OK	16:53

Data Download Date/Time: 09/30/2019 13:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASEY,COREY,M
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: CASEY,COREY
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 08/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: KLEIN

DATE/TIME OF TEST: 09/21/2019 00:04:53

Diagnostics Check	OK	23:58
Air Blank	0.000	23:59
Control Test	0.080	23:59
Air Blank	0.000	00:00
Subject Sample # 1	0.091	00:00
Breath Volume	2.863	00:00
Air Blank	0.000	00:01
Air Blank	0.000	00:02
Subject Sample # 2	0.089	00:03
Breath Volume	3.292	00:03
Air Blank	0.000	00:03
Control Test	0.080	00:04
Air Blank	0.000	00:04
Diagnostics Check	OK	00:04

Data Download Date/Time: 09/30/2019 13:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASEY,COREY,M
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: BUTLER,ADAM
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 08/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: NEIHOUSE

DATE/TIME OF TEST: 09/21/2019 00:58:19

Diagnostics Check	OK	00:50
Air Blank	0.000	00:51
Control Test	0.081	00:51
Air Blank	0.000	00:52
Subject Sample # 1	0.146	00:53
Breath Volume	1.585	00:53
Air Blank	0.000	00:53
Air Blank	0.000	00:55
Subject Sample # 2	0.146	00:56
Breath Volume	2.812	00:56
Air Blank	0.000	00:57
Control Test	0.080	00:57
Air Blank	0.000	00:58
Diagnostics Check	OK	00:58

Data Download Date/Time: 09/30/2019 13:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,T,N
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: WHITE,T
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 08/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: VOELLER

DATE/TIME OF TEST: 09/21/2019 01:28:07

Diagnostics Check	OK	01:21
Air Blank	0.000	01:22
Control Test	0.080	01:22
Air Blank	0.000	01:23
Subject Sample # 1	0.197	01:23
Breath Volume	2.343	01:23
Air Blank	0.000	01:24
Air Blank	0.000	01:26
Subject Sample # 2	0.193	01:26
Breath Volume	1.304	01:26
Air Blank	0.000	01:27
Control Test	0.079	01:27
Air Blank	0.000	01:28
Diagnostics Check	OK	01:28

Data Download Date/Time: 09/30/2019 13:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,T,N
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: AIANI,J
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 08/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: LOVERIDGE

DATE/TIME OF TEST: 09/21/2019 02:55:57

Diagnostics Check	OK	02:49
Air Blank	0.000	02:50
Control Test	0.080	02:50
Air Blank	0.000	02:51
Subject Sample # 1	0.232	02:51
Breath Volume	2.222	02:51
Air Blank	0.000	02:52
Air Blank	0.000	02:53
Subject Sample # 2	0.232	02:54
Breath Volume	3.535	02:54
Air Blank	0.000	02:55
Control Test	0.080	02:55
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

Data Download Date/Time: 09/30/2019 13:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,ANTHONY,M
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: CROSBY,AUSTIN
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 08/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: MCMASTER

DATE/TIME OF TEST: 09/28/2019 04:30:25

Diagnostics Check	OK	04:17	
Air Blank	0.000	04:17	
Control Test	0.080	04:18	
Air Blank	0.000	04:18	
Subject Sample	# 1 0.114	04:21	Slope Not Level
Breath Volume	1.414	04:21	
Air Blank	0.000	04:22	
Air Blank	0.000	04:24	
Subject Sample	# 2 0.136	04:24	
Breath Volume	2.597	04:24	
Air Blank	0.000	04:25	
Air Blank	0.000	04:27	
Subject Sample	# 3 0.129	04:28	
Breath Volume	2.742	04:28	
Air Blank	0.000	04:29	
Control Test	0.079	04:29	
Air Blank	0.000	04:30	
Diagnostics Check	OK	04:30	

Data Download Date/Time: 09/30/2019 13:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEONARD,BRIAN,J
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: KRUKOSKI,KEVIN
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 09/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 16:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:27

SUBJECT'S LAST NAME: LIVSEY

DATE/TIME OF TEST: 09/30/2019 17:20:45

Diagnostics Check	OK	17:10
Air Blank	0.000	17:11
Control Test	0.077	17:11
Air Blank	0.000	17:11
Subject Sample # 1	0.344	17:12
Breath Volume	1.660	17:12
Air Blank	0.000	17:13
Air Blank	0.000	17:14
Subject Sample # 2	0.318	17:15
Breath Volume	2.226	17:15
Air Blank	0.000	17:16
Air Blank	0.000	17:18
Subject Sample # 3	0.324	17:18
Breath Volume	1.554	17:18
Air Blank	0.000	17:19
Control Test	0.075	17:20
Air Blank	0.000	17:20
Diagnostics Check	OK	17:20

Data Download Date/Time: 10/31/2019 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

INSTRUMENT REGISTERED TO: INDIAN HARBOUR PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGRORY, DIARMUID, J
BREATH TEST OPERATOR AGENCY: Indian Harbour Beach
ARRESTING OFFICER: COCCHIOLA, AUSTIN
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: WILCOX

DATE/TIME OF TEST: 09/06/2019 00:07:05

Diagnostics Check	OK	00:00
Air Blank	0.000	00:01
Control Test	0.081	00:01
Air Blank	0.000	00:02
Subject Sample # 1	0.000	00:02
Breath Volume	1.855	00:02
Air Blank	0.000	00:03
Air Blank	0.000	00:05
Subject Sample # 2	0.000	00:05
Breath Volume	1.593	00:05
Air Blank	0.000	00:06
Control Test	0.081	00:06
Air Blank	0.000	00:06
Diagnostics Check	OK	00:07

Data Download Date/Time: 09/13/2019 05:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

INSTRUMENT REGISTERED TO: INDIAN HARBOUR PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OAKS,KEVIN,L
BREATH TEST OPERATOR AGENCY: Indian Harbour Beach
ARRESTING OFFICER: BERGERON,J
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 09:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:00

SUBJECT'S LAST NAME: DAILEY

DATE/TIME OF TEST: 09/06/2019 10:45:38

Diagnostics Check	OK	10:44	
Air Blank	0.000	10:44	
Control Test	0.216	10:44	Control Outside Tolerance
Air Blank	0.000	10:45	

Data Download Date/Time: 09/13/2019 05:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

INSTRUMENT REGISTERED TO: INDIAN HARBOUR PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OAKS,KEVIN,L
BREATH TEST OPERATOR AGENCY: Indian Harbour Beach
ARRESTING OFFICER: BERGERON,J
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 09:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:00

SUBJECT'S LAST NAME: DAILEY

DATE/TIME OF TEST: 09/06/2019 10:54:11

Diagnostics Check	OK	10:47
Air Blank	0.000	10:48
Control Test	0.080	10:48
Air Blank	0.000	10:49
Subject Sample # 1	0.259	10:49
Breath Volume	2.429	10:49
Air Blank	0.000	10:50
Air Blank	0.000	10:52
Subject Sample # 2	0.245	10:52
Breath Volume	2.375	10:52
Air Blank	0.000	10:53
Control Test	0.079	10:53
Air Blank	0.000	10:54
Diagnostics Check	OK	10:54

Data Download Date/Time: 09/13/2019 05:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JENKINS,CHRISTINA,E
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: JENKINS,CHRISTINA
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 08/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A3
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: LITTLE

DATE/TIME OF TEST: 09/01/2019 22:44:00

Diagnostics Check	OK	22:37
Air Blank	0.000	22:38
Control Test	0.082	22:38
Air Blank	0.000	22:39
Subject Sample # 1	0.000	22:39
Breath Volume	1.605	22:39
Air Blank	0.000	22:40
Air Blank	0.000	22:42
Subject Sample # 2	0.000	22:42
Breath Volume	1.570	22:42
Air Blank	0.000	22:43
Control Test	0.081	22:43
Air Blank	0.000	22:43
Diagnostics Check	OK	22:44

Data Download Date/Time: 09/26/2019 00:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,MARCUS,H
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BALGO,HARRY
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 08/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A3
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 09/03/2019 04:51:21

Diagnostics Check	OK	04:44
Air Blank	0.000	04:45
Control Test	0.081	04:45
Air Blank	0.000	04:45
Subject Sample # 1	0.220	04:46
Breath Volume	2.839	04:46
Air Blank	0.000	04:47
Air Blank	0.000	04:49
Subject Sample # 2	0.220	04:49
Breath Volume	3.027	04:49
Air Blank	0.000	04:50
Control Test	0.080	04:50
Air Blank	0.000	04:51
Diagnostics Check	OK	04:51

Data Download Date/Time: 09/26/2019 00:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINSTEAD,JUSTIN,M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: WINSTEAD,JUSTIN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 08/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A3
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: CARACCIOLO

DATE/TIME OF TEST: 09/05/2019 00:53:42

Diagnostics Check	OK	00:39	
Air Blank	0.000	00:39	
Control Test	0.081	00:40	
Air Blank	0.000	00:40	
Subject Sample	# 1 0.115	00:43	Volume Not Met
Breath Volume	0.246	00:43	
Air Blank	0.000	00:44	
Air Blank	0.000	00:46	
Subject Sample	# 2 0.135	00:47	
Breath Volume	1.972	00:47	
Air Blank	0.000	00:48	
Air Blank	0.000	00:50	
Subject Sample	# 3 0.141	00:51	
Breath Volume	2.660	00:51	
Air Blank	0.000	00:52	
Control Test	0.077	00:53	
Air Blank	0.000	00:53	
Diagnostics Check	OK	00:53	

Data Download Date/Time: 09/26/2019 00:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,FREDERICK,A
BREATH TEST OPERATOR AGENCY: Kennedy Space Center
ARRESTING OFFICER: BUSHNELL,B
ARRESTING OFFICER AGENCY: OTHER Fed
LAST AGENCY INSPECTION: 08/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 13:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:17

SUBJECT'S LAST NAME: COOLER

DATE/TIME OF TEST: 09/07/2019 15:26:09

Diagnostics Check	OK	15:19
Air Blank	0.000	15:20
Control Test	0.082	15:20
Air Blank	0.000	15:21
Subject Sample # 1	0.129	15:21
Breath Volume	2.156	15:21
Air Blank	0.000	15:22
Air Blank	0.000	15:24
Subject Sample # 2	0.133	15:24
Breath Volume	1.812	15:24
Air Blank	0.000	15:25
Control Test	0.080	15:25
Air Blank	0.000	15:26
Diagnostics Check	OK	15:26

Data Download Date/Time: 09/26/2019 00:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,FREDERICK,A
BREATH TEST OPERATOR AGENCY: Kennedy Space Center
ARRESTING OFFICER: POWERS,STEVEN
ARRESTING OFFICER AGENCY: OTHER Fed
LAST AGENCY INSPECTION: 08/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 14:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:02

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 09/07/2019 15:43:03

Diagnostics Check	OK	15:36
Air Blank	0.000	15:37
Control Test	0.081	15:37
Air Blank	0.000	15:38
Subject Sample # 1	0.106	15:38
Breath Volume	3.835	15:38
Air Blank	0.000	15:39
Air Blank	0.000	15:41
Subject Sample # 2	0.104	15:41
Breath Volume	3.851	15:41
Air Blank	0.000	15:42
Control Test	0.080	15:42
Air Blank	0.000	15:42
Diagnostics Check	OK	15:43

Data Download Date/Time: 09/26/2019 00:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANLEY,ROBERT,W
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: GRAY,CONNOR
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 08/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: WILLIAMS III

DATE/TIME OF TEST: 09/09/2019 20:57:07

Diagnostics Check	OK	20:51
Air Blank	0.000	20:51
Control Test	0.081	20:51
Air Blank	0.000	20:52
Subject Sample # 1	0.293	20:52
Breath Volume	2.351	20:52
Air Blank	0.000	20:53
Air Blank	0.000	20:55
Subject Sample # 2	0.280	20:55
Breath Volume	1.628	20:55
Air Blank	0.000	20:56
Control Test	0.079	20:56
Air Blank	0.000	20:57
Diagnostics Check	OK	20:57

Data Download Date/Time: 09/26/2019 00:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSON,JASON,A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: ROSA,EMILO
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 08/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 09/12/2019 02:34:59

Diagnostics Check	OK	02:22	
Air Blank	0.000	02:23	
Control Test	0.081	02:23	
Air Blank	0.000	02:24	
Subject Sample	# 1 0.053	02:27	Volume Not Met
Breath Volume	1.343	02:27	
Air Blank	0.000	02:28	
Air Blank	0.000	02:29	
Subject Sample	# 2 0.055	02:30	
Breath Volume	1.828	02:30	
Air Blank	0.000	02:31	
Air Blank	0.000	02:32	
Subject Sample	# 3 0.055	02:33	
Breath Volume	1.921	02:33	
Air Blank	0.000	02:34	
Control Test	0.078	02:34	
Air Blank	0.000	02:34	
Diagnostics Check	OK	02:34	

Data Download Date/Time: 09/26/2019 00:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANLEY,ROBERT,W
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: APPLGATE,KAITLYN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 08/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: REEDER

DATE/TIME OF TEST: 09/14/2019 20:45:39

Diagnostics Check	OK	20:38
Air Blank	0.000	20:38
Control Test	0.080	20:39
Air Blank	0.000	20:39
Subject Sample # 1	0.178	20:41
Breath Volume	1.238	20:41
Air Blank	0.000	20:42
Air Blank	0.000	20:43
Subject Sample # 2	0.167	20:44
Breath Volume	2.050	20:44
Air Blank	0.000	20:44
Control Test	0.079	20:45
Air Blank	0.000	20:45
Diagnostics Check	OK	20:45

Data Download Date/Time: 09/26/2019 00:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANLEY,ROBERT,W
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: HANNON,MITCHELL
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 08/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: BUNCH

DATE/TIME OF TEST: 09/15/2019 03:00:45

Diagnostics Check	OK	02:54	
Air Blank	0.000	02:54	
Control Test	0.082	02:54	
Air Blank	0.000	02:55	
Subject Sample	# 1 0.238	02:55	
Breath Volume	1.832	02:55	
Air Blank	0.000	02:56	
Air Blank	0.000	02:58	
Subject Sample	# 2 0.236	03:00	
Breath Volume	1.625	03:00	
Air Blank		03:00	RFI Detect
Air Blank		03:00	RFI Detect

Data Download Date/Time: 09/26/2019 00:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANLEY,ROBERT,W
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: HANNON,MITCHELL
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 08/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: BUNCH

DATE/TIME OF TEST: 09/15/2019 03:06:33

Diagnostics Check	OK	03:02	
Air Blank	0.000	03:03	
Control Test	0.078	03:03	
Air Blank	0.000	03:03	
Subject Sample	# 1 0.238	03:05	
Breath Volume	1.171	03:05	
Air Blank		03:05	Purge Fail
Air Blank	0.000	03:06	

Data Download Date/Time: 09/26/2019 00:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANLEY,ROBERT,W
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: HANNON,MITCHELL
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 08/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: BUNCH

DATE/TIME OF TEST: 09/15/2019 03:21:56

Diagnostics Check	OK	03:09	
Air Blank	0.000	03:09	
Control Test	0.081	03:09	
Air Blank	0.000	03:10	
Subject Sample	# 1 0.176	03:13	Volume Not Met
Breath Volume	0.328	03:13	
Air Blank	0.000	03:14	
Air Blank	0.000	03:16	
Subject Sample	# 2 0.227	03:16	
Breath Volume	1.421	03:16	
Air Blank	0.000	03:17	
Air Blank	0.000	03:19	
Subject Sample	# 3 0.217	03:20	
Breath Volume	1.515	03:20	
Air Blank	0.000	03:21	
Control Test	0.076	03:21	
Air Blank	0.000	03:21	
Diagnostics Check	OK	03:21	

Data Download Date/Time: 09/26/2019 00:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,ROBERT,F
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: WAGNER,ROBERT
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 08/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: EMOND

DATE/TIME OF TEST: 09/20/2019 02:49:00

Diagnostics Check	OK	02:42
Air Blank	0.000	02:43
Control Test	0.081	02:43
Air Blank	0.000	02:44
Subject Sample # 1	0.124	02:44
Breath Volume	1.988	02:44
Air Blank	0.000	02:45
Air Blank	0.000	02:46
Subject Sample # 2	0.126	02:47
Breath Volume	2.851	02:47
Air Blank	0.000	02:48
Control Test	0.081	02:48
Air Blank	0.000	02:48
Diagnostics Check	OK	02:49

Data Download Date/Time: 09/26/2019 00:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,ROBERT,F
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MANLEY,ROBERT
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 08/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: ST LOUIS

DATE/TIME OF TEST: 09/20/2019 03:04:45

Diagnostics Check	OK	02:55
Air Blank	0.000	02:56
Control Test	0.082	02:56
Air Blank	0.000	02:57
Subject Sample # 1	0.125	02:59
Breath Volume	1.167	02:59
Air Blank	0.000	02:59
Air Blank	0.000	03:01
Subject Sample # 2	0.125	03:03
Breath Volume	1.679	03:03
Air Blank	0.000	03:03
Control Test	0.080	03:04
Air Blank	0.000	03:04
Diagnostics Check	OK	03:04

Data Download Date/Time: 09/26/2019 00:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,ROBERT,F
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: DAUGIRDAS,CHRISTIAN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 08/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: BRINSON

DATE/TIME OF TEST: 09/21/2019 22:16:59

Diagnostics Check	OK	22:10
Air Blank	0.000	22:10
Control Test	0.082	22:11
Air Blank	0.000	22:11
Subject Sample # 1	0.138	22:12
Breath Volume	1.960	22:12
Air Blank	0.000	22:13
Air Blank	0.000	22:14
Subject Sample # 2	0.137	22:15
Breath Volume	1.718	22:15
Air Blank	0.000	22:16
Control Test	0.081	22:16
Air Blank	0.000	22:16
Diagnostics Check	OK	22:16

Data Download Date/Time: 09/26/2019 00:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINSTEAD,JUSTIN,M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: HANNON,MITCHELL
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 08/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: COLEMAN

DATE/TIME OF TEST: 09/24/2019 23:42:34

Diagnostics Check	OK	23:39	
Air Blank	0.000	23:40	
Control Test	0.082	23:40	
Air Blank	0.000	23:41	
Subject Sample	# 1	23:41	Subject Test Refused
Breath Volume	0.000	23:41	
Air Blank	0.000	23:41	
Control Test	0.081	23:41	
Air Blank	0.000	23:42	
Diagnostics Check	OK	23:42	

Data Download Date/Time: 09/26/2019 00:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 08/28/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17919080A1

DRY GAS EXPIRY DATE: 08/05/2021

VIOLATION CODE: DUI

ARREST TIME: 23:45

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: HAMPTON

DATE/TIME OF TEST: 09/01/2019 00:04:48

Diagnostics Check	OK	23:58
Air Blank	0.000	23:59
Control Test	0.080	23:59
Air Blank	0.000	00:00
Subject Sample # 1	0.139	00:00
Breath Volume	1.300	00:00
Air Blank	0.000	00:00
Air Blank	0.000	00:02
Subject Sample # 2	0.156	00:03
Breath Volume	1.437	00:03
Air Blank	0.000	00:03
Control Test	0.079	00:04
Air Blank	0.000	00:04
Diagnostics Check	OK	00:04

Data Download Date/Time: 09/04/2019 14:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 08/28/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17919080A1

DRY GAS EXPIRY DATE: 08/05/2021

VIOLATION CODE: DUI

ARREST TIME: 03:20

GENDER: Female

OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: BREUER

DATE/TIME OF TEST: 09/01/2019 03:33:10

Diagnostics Check	OK	03:27
Air Blank	0.000	03:27
Control Test	0.079	03:27
Air Blank	0.000	03:28
Subject Sample # 1	0.150	03:28
Breath Volume	1.882	03:28
Air Blank	0.000	03:29
Air Blank	0.000	03:31
Subject Sample # 2	0.144	03:31
Breath Volume	1.625	03:31
Air Blank	0.000	03:32
Control Test	0.078	03:32
Air Blank	0.000	03:33
Diagnostics Check	OK	03:33

Data Download Date/Time: 09/04/2019 14:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 08/28/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17919080A1

DRY GAS EXPIRY DATE: 08/05/2021

VIOLATION CODE: DUI

ARREST TIME: 21:04

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:53

SUBJECT'S LAST NAME: TEAGUE

DATE/TIME OF TEST: 09/06/2019 21:19:09

Diagnostics Check	OK	21:13
Air Blank	0.000	21:13
Control Test	0.079	21:13
Air Blank	0.000	21:14
Subject Sample # 1	0.130	21:14
Breath Volume	2.421	21:14
Air Blank	0.000	21:15
Air Blank	0.000	21:17
Subject Sample # 2	0.134	21:17
Breath Volume	1.851	21:17
Air Blank	0.000	21:18
Control Test	0.081	21:18
Air Blank	0.000	21:19
Diagnostics Check	OK	21:19

Data Download Date/Time: 10/04/2019 00:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN - VOL SAMPLE

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 08/28/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17919080A1

DRY GAS EXPIRY DATE: 08/05/2021

VIOLATION CODE: OTHER

ARREST TIME: 00:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: BARRETTE

DATE/TIME OF TEST: 09/08/2019 02:27:34

Diagnostics Check	OK	02:21
Air Blank	0.000	02:21
Control Test	0.079	02:22
Air Blank	0.000	02:22
Subject Sample # 1	0.049	02:23
Breath Volume	2.062	02:23
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample # 2	0.046	02:26
Breath Volume	1.433	02:26
Air Blank	0.000	02:26
Control Test	0.078	02:26
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

Data Download Date/Time: 10/04/2019 00:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 08/28/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17919080A1

DRY GAS EXPIRY DATE: 08/05/2021

VIOLATION CODE: DUI

ARREST TIME: 21:47

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: CONRAD

DATE/TIME OF TEST: 09/08/2019 22:02:34

Diagnostics Check	OK	21:56
Air Blank	0.000	21:57
Control Test	0.079	21:57
Air Blank	0.000	21:57
Subject Sample # 1	0.094	21:58
Breath Volume	3.617	21:58
Air Blank	0.000	21:58
Air Blank	0.000	22:00
Subject Sample # 2	0.094	22:01
Breath Volume	3.597	22:01
Air Blank	0.000	22:01
Control Test	0.079	22:01
Air Blank	0.000	22:02
Diagnostics Check	OK	22:02

Data Download Date/Time: 10/04/2019 00:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 08/28/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17919080A1

DRY GAS EXPIRY DATE: 08/05/2021

VIOLATION CODE: DUI

ARREST TIME: 02:23

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: MYERS

DATE/TIME OF TEST: 09/14/2019 02:36:53

Diagnostics Check	OK	02:30
Air Blank	0.000	02:31
Control Test	0.080	02:31
Air Blank	0.000	02:31
Subject Sample # 1	0.093	02:32
Breath Volume	1.410	02:32
Air Blank	0.000	02:32
Air Blank	0.000	02:34
Subject Sample # 2	0.092	02:35
Breath Volume	1.492	02:35
Air Blank	0.000	02:35
Control Test	0.079	02:36
Air Blank	0.000	02:36
Diagnostics Check	OK	02:36

Data Download Date/Time: 10/04/2019 00:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 08/28/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17919080A1

DRY GAS EXPIRY DATE: 08/05/2021

VIOLATION CODE: DUI

ARREST TIME: 00:25

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: BLACKETT

DATE/TIME OF TEST: 09/15/2019 00:35:07

Diagnostics Check	OK	00:34	
Air Blank	0.000	00:34	
		00:34	Sequence Aborted
Air Blank	0.000	00:35	

Data Download Date/Time: 10/04/2019 00:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 08/28/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17919080A1

DRY GAS EXPIRY DATE: 08/05/2021

VIOLATION CODE: DUI

ARREST TIME: 00:25

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: BLACKETT

DATE/TIME OF TEST: 09/15/2019 01:01:39

Diagnostics Check	OK	00:55
Air Blank	0.000	00:56
Control Test	0.079	00:56
Air Blank	0.000	00:56
Subject Sample # 1	0.097	00:57
Breath Volume	1.835	00:57
Air Blank	0.000	00:57
Air Blank	0.000	00:59
Subject Sample # 2	0.096	01:00
Breath Volume	2.304	01:00
Air Blank	0.000	01:00
Control Test	0.078	01:01
Air Blank	0.000	01:01
Diagnostics Check	OK	01:01

Data Download Date/Time: 10/04/2019 00:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 08/28/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17919080A1

DRY GAS EXPIRY DATE: 08/05/2021

VIOLATION CODE: DUI

ARREST TIME: 22:50

GENDER: Female

OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: HORTON

DATE/TIME OF TEST: 09/15/2019 23:04:02

Diagnostics Check	OK	22:57
Air Blank	0.000	22:58
Control Test	0.079	22:58
Air Blank	0.000	22:58
Subject Sample # 1	0.000	22:59
Breath Volume	2.281	22:59
Air Blank	0.000	22:59
Air Blank	0.000	23:01
Subject Sample # 2	0.000	23:02
Breath Volume	2.042	23:02
Air Blank	0.000	23:03
Control Test	0.079	23:03
Air Blank	0.000	23:03
Diagnostics Check	OK	23:04

Data Download Date/Time: 10/04/2019 00:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 08/28/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17919080A1

DRY GAS EXPIRY DATE: 08/05/2021

VIOLATION CODE: DUI

ARREST TIME: 23:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: BARTHOLOMEW

DATE/TIME OF TEST: 09/21/2019 23:57:41

Diagnostics Check	OK	23:51
Air Blank	0.000	23:51
Control Test	0.079	23:51
Air Blank	0.000	23:52
Subject Sample # 1	0.210	23:53
Breath Volume	2.203	23:53
Air Blank	0.000	23:53
Air Blank	0.000	23:55
Subject Sample # 2	0.201	23:56
Breath Volume	1.218	23:56
Air Blank	0.000	23:56
Control Test	0.079	23:57
Air Blank	0.000	23:57
Diagnostics Check	OK	23:57

Data Download Date/Time: 10/04/2019 00:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILL,SEAN,B
BREATH TEST OPERATOR AGENCY: Sanford PD
ARRESTING OFFICER: HILL,SEAN
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 19:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: HARRINGTON

DATE/TIME OF TEST: 09/01/2019 20:02:18

Diagnostics Check	OK	19:55
Air Blank	0.000	19:56
Control Test	0.081	19:56
Air Blank	0.000	19:57
Subject Sample # 1	0.067	19:57
Breath Volume	1.777	19:57
Air Blank	0.000	19:58
Air Blank	0.000	20:00
Subject Sample # 2	0.063	20:00
Breath Volume	1.277	20:00
Air Blank	0.000	20:01
Control Test	0.081	20:01
Air Blank	0.000	20:02
Diagnostics Check	OK	20:02

Data Download Date/Time: 10/01/2019 00:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOHRI,BRENDAN,R
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: BADALL,ALEXANDER
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: REGER

DATE/TIME OF TEST: 09/02/2019 01:01:22

Diagnostics Check	OK	00:54
Air Blank	0.000	00:55
Control Test	0.080	00:55
Air Blank	0.000	00:55
Subject Sample # 1	0.213	00:56
Breath Volume	1.878	00:56
Air Blank	0.000	00:57
Air Blank	0.000	00:59
Subject Sample # 2	0.218	00:59
Breath Volume	2.566	00:59
Air Blank	0.000	01:00
Control Test	0.080	01:00
Air Blank	0.000	01:01
Diagnostics Check	OK	01:01

Data Download Date/Time: 10/01/2019 00:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTANA,ABEL,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: ANDREWS,GARY
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 09/02/2019 04:18:48

Diagnostics Check	OK	04:12
Air Blank	0.000	04:12
Control Test	0.081	04:13
Air Blank	0.000	04:13
Subject Sample # 1	0.100	04:14
Breath Volume	2.863	04:14
Air Blank	0.000	04:14
Air Blank	0.000	04:16
Subject Sample # 2	0.105	04:17
Breath Volume	3.230	04:17
Air Blank	0.000	04:17
Control Test	0.081	04:18
Air Blank	0.000	04:18
Diagnostics Check	OK	04:18

Data Download Date/Time: 10/01/2019 00:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOSUE,
BREATH TEST OPERATOR AGENCY: Altamonte Springs PD
ARRESTING OFFICER: VENTURA,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: VALLE

DATE/TIME OF TEST: 09/03/2019 01:28:33

Diagnostics Check	OK	01:21
Air Blank	0.000	01:22
Control Test	0.081	01:22
Air Blank	0.000	01:23
Subject Sample # 1	0.204	01:23
Breath Volume	1.742	01:23
Air Blank	0.000	01:24
Air Blank	0.000	01:26
Subject Sample # 2	0.189	01:27
Breath Volume	1.164	01:27
Air Blank	0.000	01:27
Control Test	0.081	01:27
Air Blank	0.000	01:28
Diagnostics Check	OK	01:28

Data Download Date/Time: 10/01/2019 00:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOSUE,
BREATH TEST OPERATOR AGENCY: Altamonte Springs PD
ARRESTING OFFICER: KINKOPF,TIM
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: ESPOSITO

DATE/TIME OF TEST: 09/03/2019 02:03:37

Diagnostics Check	OK	01:56
Air Blank	0.000	01:57
Control Test	0.081	01:57
Air Blank	0.000	01:57
Subject Sample # 1	0.208	01:59
Breath Volume	2.082	01:59
Air Blank	0.000	01:59
Air Blank	0.000	02:01
Subject Sample # 2	0.207	02:02
Breath Volume	1.570	02:02
Air Blank	0.000	02:02
Control Test	0.080	02:03
Air Blank	0.000	02:03
Diagnostics Check	OK	02:03

Data Download Date/Time: 10/01/2019 00:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOSUE,
BREATH TEST OPERATOR AGENCY: Altamonte Springs PD
ARRESTING OFFICER: MASSIP,LEONARDO
ARRESTING OFFICER AGENCY: Oviedo PD
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: LAGUERRE

DATE/TIME OF TEST: 09/04/2019 04:11:50

Diagnostics Check	OK	04:05
Air Blank	0.000	04:05
Control Test	0.081	04:05
Air Blank	0.000	04:06
Subject Sample # 1	0.000	04:07
Breath Volume	1.875	04:07
Air Blank	0.000	04:07
Air Blank	0.000	04:09
Subject Sample # 2	0.000	04:10
Breath Volume	2.398	04:10
Air Blank	0.000	04:10
Control Test	0.081	04:11
Air Blank	0.000	04:11
Diagnostics Check	OK	04:11

Data Download Date/Time: 10/01/2019 00:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GARCIA, YOEL,

BREATH TEST OPERATOR AGENCY: FHP Troop D

ARRESTING OFFICER: PONDER, CHRISTOPHER

ARRESTING OFFICER AGENCY: FHP Troop D

LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 915310

DRY GAS EXPIRY DATE: 03/07/2021

VIOLATION CODE: DUI

ARREST TIME: 00:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: DEJESUS

DATE/TIME OF TEST: 09/05/2019 01:25:17

Diagnostics Check	OK	01:22	
Air Blank	0.000	01:23	
Control Test	0.082	01:23	
Air Blank	0.000	01:24	
Subject Sample	# 1	01:24	Sequence Aborted
Breath Volume	0.000	01:24	
Air Blank	0.000	01:25	

Data Download Date/Time: 10/01/2019 00:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOHRI,BRENDAN,R
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: LOHRI,BRENDAN
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: LOMIDZE

DATE/TIME OF TEST: 09/05/2019 04:03:15

Diagnostics Check	OK	03:56
Air Blank	0.000	03:57
Control Test	0.082	03:57
Air Blank	0.000	03:57
Subject Sample # 1	0.151	03:58
Breath Volume	2.250	03:58
Air Blank	0.000	03:59
Air Blank	0.000	04:01
Subject Sample # 2	0.144	04:01
Breath Volume	2.507	04:01
Air Blank	0.000	04:02
Control Test	0.080	04:02
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time: 10/01/2019 00:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOHRI,BRENDAN,R
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: DEL CASTILLO,BRET
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: LANGDON

DATE/TIME OF TEST: 09/05/2019 04:22:39

Diagnostics Check	OK	04:16
Air Blank	0.000	04:16
Control Test	0.081	04:17
Air Blank	0.000	04:17
Subject Sample # 1	0.193	04:18
Breath Volume	2.097	04:18
Air Blank	0.000	04:18
Air Blank	0.000	04:20
Subject Sample # 2	0.197	04:21
Breath Volume	1.906	04:21
Air Blank	0.000	04:21
Control Test	0.081	04:22
Air Blank	0.000	04:22
Diagnostics Check	OK	04:22

Data Download Date/Time: 10/01/2019 00:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOSUE,
BREATH TEST OPERATOR AGENCY: Altamonte Springs PD
ARRESTING OFFICER: TORRES,JOSUE
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 04:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: BASSARAGH

DATE/TIME OF TEST: 09/07/2019 05:53:51

Diagnostics Check	OK	05:47
Air Blank	0.000	05:47
Control Test	0.081	05:48
Air Blank	0.000	05:48
Subject Sample # 1	0.128	05:49
Breath Volume	1.902	05:49
Air Blank	0.000	05:49
Air Blank	0.000	05:51
Subject Sample # 2	0.128	05:52
Breath Volume	1.445	05:52
Air Blank	0.000	05:52
Control Test	0.081	05:53
Air Blank	0.000	05:53
Diagnostics Check	OK	05:53

Data Download Date/Time: 10/01/2019 00:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: HOLLAND,TROY
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: SCHIOPPA

DATE/TIME OF TEST: 09/07/2019 22:30:25

Diagnostics Check	OK	22:23
Air Blank	0.000	22:23
Control Test	0.080	22:23
Air Blank	0.000	22:24
Subject Sample # 1	0.179	22:25
Breath Volume	1.886	22:25
Air Blank	0.000	22:26
Air Blank	0.000	22:27
Subject Sample # 2	0.178	22:28
Breath Volume	1.882	22:28
Air Blank	0.000	22:29
Control Test	0.081	22:29
Air Blank	0.000	22:30
Diagnostics Check	OK	22:30

Data Download Date/Time: 10/01/2019 00:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: CASSATO,KAYLA
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 04:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: OYEBOLA

DATE/TIME OF TEST: 09/10/2019 05:09:33

Diagnostics Check	OK	05:03
Air Blank	0.000	05:03
Control Test	0.081	05:04
Air Blank	0.000	05:04
Subject Sample # 1	0.164	05:05
Breath Volume	1.605	05:05
Air Blank	0.000	05:05
Air Blank	0.000	05:07
Subject Sample # 2	0.167	05:08
Breath Volume	2.058	05:08
Air Blank	0.000	05:08
Control Test	0.080	05:08
Air Blank	0.000	05:09
Diagnostics Check	OK	05:09

Data Download Date/Time: 10/01/2019 00:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRENCH,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Sanford PD
ARRESTING OFFICER: FRENCH,CHRISTOPHER
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 14:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:20

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 09/11/2019 15:33:59

Diagnostics Check	OK	15:24
Air Blank	0.000	15:25
Control Test	0.082	15:25
Air Blank	0.000	15:26
Subject Sample # 1	0.000	15:28
Breath Volume	1.109	15:28
Air Blank	0.000	15:28
Air Blank	0.000	15:30
Subject Sample # 2	0.000	15:32
Breath Volume	1.167	15:32
Air Blank	0.000	15:33
Control Test	0.082	15:33
Air Blank	0.000	15:33
Diagnostics Check	OK	15:33

Data Download Date/Time: 10/01/2019 00:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: MASSIP,LEONARDO
ARRESTING OFFICER AGENCY: Oviedo PD
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: GUILD

DATE/TIME OF TEST: 09/13/2019 02:29:18

Diagnostics Check	OK	02:22
Air Blank	0.000	02:23
Control Test	0.081	02:23
Air Blank	0.000	02:24
Subject Sample # 1	0.173	02:24
Breath Volume	1.312	02:24
Air Blank	0.000	02:25
Air Blank	0.000	02:27
Subject Sample # 2	0.168	02:27
Breath Volume	1.894	02:27
Air Blank	0.000	02:28
Control Test	0.081	02:28
Air Blank	0.000	02:29
Diagnostics Check	OK	02:29

Data Download Date/Time: 10/01/2019 00:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: CANNADAY,ZACHARY
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: HUTTON

DATE/TIME OF TEST: 09/14/2019 04:32:20

Diagnostics Check	OK	04:25
Air Blank	0.000	04:25
Control Test	0.081	04:26
Air Blank	0.000	04:26
Subject Sample # 1	0.169	04:27
Breath Volume	3.738	04:27
Air Blank	0.000	04:27
Air Blank	0.000	04:29
Subject Sample # 2	0.173	04:30
Breath Volume	2.492	04:30
Air Blank	0.000	04:31
Control Test	0.082	04:31
Air Blank	0.000	04:32
Diagnostics Check	OK	04:32

Data Download Date/Time: 10/01/2019 00:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMENS,ADAM,J
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: CAFFEY,Q
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 06:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:42

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 09/14/2019 08:38:57

Diagnostics Check	OK	08:29
Air Blank	0.000	08:30
Control Test	0.080	08:30
Air Blank	0.000	08:31
Subject Sample # 1	0.079	08:31
Breath Volume	1.703	08:31
Air Blank	0.000	08:32
Air Blank	0.000	08:34
Subject Sample # 2	0.074	08:37
Breath Volume	1.484	08:37
Air Blank	0.000	08:38
Control Test	0.080	08:38
Air Blank	0.000	08:38
Diagnostics Check	OK	08:38

Data Download Date/Time: 10/01/2019 00:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: DAVIS,SPENCER
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 09/15/2019 00:44:25

Diagnostics Check	OK	00:38
Air Blank	0.000	00:38
Control Test	0.081	00:39
Air Blank	0.000	00:39
Subject Sample # 1	0.000	00:40
Breath Volume	3.253	00:40
Air Blank	0.000	00:40
Air Blank	0.000	00:42
Subject Sample # 2	0.000	00:43
Breath Volume	1.796	00:43
Air Blank	0.000	00:43
Control Test	0.079	00:43
Air Blank	0.000	00:44
Diagnostics Check	OK	00:44

Data Download Date/Time: 10/01/2019 00:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: LANE,DAPHNE
ARRESTING OFFICER AGENCY: Longwood PD
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:39

SUBJECT'S LAST NAME: ARCARIO

DATE/TIME OF TEST: 09/15/2019 05:07:38

Diagnostics Check	OK	05:01
Air Blank	0.000	05:01
Control Test	0.081	05:01
Air Blank	0.000	05:02
Subject Sample # 1	0.115	05:02
Breath Volume	1.496	05:02
Air Blank	0.000	05:03
Air Blank	0.000	05:05
Subject Sample # 2	0.128	05:06
Breath Volume	3.289	05:06
Air Blank	0.000	05:06
Control Test	0.080	05:07
Air Blank	0.000	05:07
Diagnostics Check	OK	05:07

Data Download Date/Time: 10/01/2019 00:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: NORRIS,MARK
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 04:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: VELAZQUEZ

DATE/TIME OF TEST: 09/15/2019 05:22:45

Diagnostics Check	OK	05:15
Air Blank	0.000	05:15
Control Test	0.080	05:15
Air Blank	0.000	05:16
Subject Sample # 1	0.203	05:18
Breath Volume	1.164	05:18
Air Blank	0.000	05:18
Air Blank	0.000	05:20
Subject Sample # 2	0.199	05:21
Breath Volume	1.378	05:21
Air Blank	0.000	05:21
Control Test	0.080	05:22
Air Blank	0.000	05:22
Diagnostics Check	OK	05:22

Data Download Date/Time: 10/01/2019 00:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINKOPF,TIM,E
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: FLANAGAN,BRENT
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 20:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:51

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 09/17/2019 21:24:09

Diagnostics Check	OK	21:14
Air Blank	0.000	21:14
Control Test	0.081	21:15
Air Blank	0.000	21:15
Subject Sample # 1	0.264	21:16
Breath Volume	1.253	21:16
Air Blank	0.000	21:16
Air Blank	0.000	21:18
Subject Sample # 2	0.234	21:19
Breath Volume	1.511	21:19
Air Blank	0.000	21:19
Air Blank	0.000	21:21
Subject Sample # 3	0.234	21:22
Breath Volume	1.960	21:22
Air Blank	0.000	21:23
Control Test	0.081	21:23
Air Blank	0.000	21:24
Diagnostics Check	OK	21:24

Data Download Date/Time: 10/01/2019 00:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOUROUMOUSIS,GEORGE,T
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: TORRES,JOSUE
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: DICKSON

DATE/TIME OF TEST: 09/18/2019 01:24:10

Diagnostics Check	OK	01:18
Air Blank	0.000	01:18
Control Test	0.080	01:18
Air Blank	0.000	01:19
Subject Sample # 1	0.102	01:19
Breath Volume	4.593	01:19
Air Blank	0.000	01:20
Air Blank	0.000	01:22
Subject Sample # 2	0.097	01:22
Breath Volume	4.117	01:22
Air Blank	0.000	01:23
Control Test	0.078	01:23
Air Blank	0.000	01:24
Diagnostics Check	OK	01:24

Data Download Date/Time: 10/01/2019 00:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: ARGENTINE,BRITTANY
ARRESTING OFFICER AGENCY: Oviedo PD
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: ROSEMEYER

DATE/TIME OF TEST: 09/20/2019 01:54:42

Diagnostics Check	OK	01:43
Air Blank	0.000	01:44
Control Test	0.081	01:44
Air Blank	0.000	01:44
Subject Sample # 1	0.186	01:45
Breath Volume	3.875	01:45
Air Blank	0.000	01:46
Air Blank	0.000	01:48
Subject Sample # 2	0.162	01:48
Breath Volume	1.761	01:48
Air Blank	0.000	01:49
Air Blank	0.000	01:51
Subject Sample # 3	0.157	01:53
Breath Volume	2.054	01:53
Air Blank	0.000	01:53
Control Test	0.081	01:54
Air Blank	0.000	01:54
Diagnostics Check	OK	01:54

Data Download Date/Time: 10/01/2019 00:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: SMITH,BLAKE
ARRESTING OFFICER AGENCY: Oviedo PD
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: BENDER

DATE/TIME OF TEST: 09/20/2019 02:23:59

Diagnostics Check	OK	02:16
Air Blank	0.000	02:17
Control Test	0.081	02:17
Air Blank	0.000	02:18
Subject Sample # 1	0.185	02:19
Breath Volume	3.007	02:19
Air Blank	0.000	02:20
Air Blank	0.000	02:21
Subject Sample # 2	0.178	02:22
Breath Volume	1.281	02:22
Air Blank	0.000	02:23
Control Test	0.080	02:23
Air Blank	0.000	02:23
Diagnostics Check	OK	02:23

Data Download Date/Time: 10/01/2019 00:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: MASSIP,LEONARDO
ARRESTING OFFICER AGENCY: Oviedo PD
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: FRONTZ

DATE/TIME OF TEST: 09/21/2019 01:49:40

Diagnostics Check	OK	01:42
Air Blank	0.000	01:43
Control Test	0.081	01:43
Air Blank	0.000	01:43
Subject Sample # 1	0.199	01:44
Breath Volume	2.320	01:44
Air Blank	0.000	01:45
Air Blank	0.000	01:47
Subject Sample # 2	0.198	01:48
Breath Volume	1.718	01:48
Air Blank	0.000	01:48
Control Test	0.079	01:49
Air Blank	0.000	01:49
Diagnostics Check	OK	01:49

Data Download Date/Time: 10/01/2019 00:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: HOLLAND,TROY
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: VAUGHN

DATE/TIME OF TEST: 09/21/2019 02:44:59

Diagnostics Check	OK	02:38
Air Blank	0.000	02:38
Control Test	0.080	02:38
Air Blank	0.000	02:39
Subject Sample # 1	0.138	02:40
Breath Volume	3.550	02:40
Air Blank	0.000	02:40
Air Blank	0.000	02:42
Subject Sample # 2	0.137	02:43
Breath Volume	3.746	02:43
Air Blank	0.000	02:44
Control Test	0.081	02:44
Air Blank	0.000	02:44
Diagnostics Check	OK	02:44

Data Download Date/Time: 10/01/2019 00:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,L,P
BREATH TEST OPERATOR AGENCY: Lake Mary PD
ARRESTING OFFICER: FERGUSON,R
ARRESTING OFFICER AGENCY: Lake Mary PD
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 11:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:00

SUBJECT'S LAST NAME: MONTANEZ

DATE/TIME OF TEST: 09/21/2019 13:52:51

Diagnostics Check	OK	13:49	
Air Blank	0.000	13:50	
Control Test	0.081	13:50	
Air Blank	0.000	13:51	
Subject Sample	# 1	13:51	Subject Test Refused
Breath Volume	0.000	13:51	
Air Blank	0.000	13:51	
Control Test	0.082	13:52	
Air Blank	0.000	13:52	
Diagnostics Check	OK	13:52	

Data Download Date/Time: 10/01/2019 00:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINKOPF,TIM,E
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: HERNANDEZ,DAVID
ARRESTING OFFICER AGENCY: Longwood PD
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 17:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:48

SUBJECT'S LAST NAME: SCHWEICKERT

DATE/TIME OF TEST: 09/21/2019 18:17:33

Diagnostics Check	OK	18:10
Air Blank	0.000	18:10
Control Test	0.082	18:11
Air Blank	0.000	18:11
Subject Sample # 1	0.000	18:13
Breath Volume	2.804	18:13
Air Blank	0.000	18:13
Air Blank	0.000	18:15
Subject Sample # 2	0.000	18:16
Breath Volume	3.343	18:16
Air Blank	0.000	18:16
Control Test	0.081	18:16
Air Blank	0.000	18:17
Diagnostics Check	OK	18:17

Data Download Date/Time: 10/01/2019 00:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: NAPIER,JEFFREY
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: FALKNER

DATE/TIME OF TEST: 09/22/2019 00:12:23

Diagnostics Check	OK	00:06
Air Blank	0.000	00:06
Control Test	0.081	00:06
Air Blank	0.000	00:07
Subject Sample # 1	0.298	00:07
Breath Volume	3.410	00:07
Air Blank	0.000	00:08
Air Blank	0.000	00:10
Subject Sample # 2	0.297	00:10
Breath Volume	3.230	00:10
Air Blank	0.000	00:11
Control Test	0.080	00:11
Air Blank	0.000	00:12
Diagnostics Check	OK	00:12

Data Download Date/Time: 10/01/2019 00:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: STEWART,TIMOTHY
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 09/22/2019 02:39:37

Diagnostics Check	OK	02:33
Air Blank	0.000	02:33
Control Test	0.081	02:34
Air Blank	0.000	02:34
Subject Sample # 1	0.096	02:35
Breath Volume	2.570	02:35
Air Blank	0.000	02:35
Air Blank	0.000	02:37
Subject Sample # 2	0.093	02:38
Breath Volume	2.238	02:38
Air Blank	0.000	02:38
Control Test	0.081	02:39
Air Blank	0.000	02:39
Diagnostics Check	OK	02:39

Data Download Date/Time: 10/01/2019 00:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTANA,ABEL,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: QUINTANA,ABEL
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 15:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: OCAMPO

DATE/TIME OF TEST: 09/22/2019 17:09:07

Diagnostics Check	OK	17:02
Air Blank	0.000	17:03
Control Test	0.082	17:03
Air Blank	0.000	17:04
Subject Sample # 1	0.063	17:04
Breath Volume	1.164	17:04
Air Blank	0.000	17:05
Air Blank	0.000	17:07
Subject Sample # 2	0.063	17:07
Breath Volume	1.746	17:07
Air Blank	0.000	17:08
Control Test	0.082	17:08
Air Blank	0.000	17:09
Diagnostics Check	OK	17:09

Data Download Date/Time: 10/01/2019 00:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,L,P
BREATH TEST OPERATOR AGENCY: Lake Mary PD
ARRESTING OFFICER: OVERALL,J
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 08/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 14:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:56

SUBJECT'S LAST NAME: ROSE

DATE/TIME OF TEST: 09/26/2019 15:29:21

Diagnostics Check	OK	15:21
Air Blank	0.000	15:22
Control Test	0.080	15:22
Air Blank	0.000	15:23
Subject Sample # 1	0.202	15:24
Breath Volume	2.367	15:24
Air Blank	0.000	15:25
Air Blank	0.000	15:27
Subject Sample # 2	0.195	15:27
Breath Volume	2.054	15:27
Air Blank	0.000	15:28
Control Test	0.080	15:28
Air Blank	0.000	15:29
Diagnostics Check	OK	15:29

Data Download Date/Time: 10/01/2019 00:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,TROY,M
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: DEL CASTILLO,BRET
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 09/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: KIM

DATE/TIME OF TEST: 09/27/2019 23:31:05

Diagnostics Check	OK	23:25
Air Blank	0.000	23:25
Control Test	0.081	23:25
Air Blank	0.000	23:26
Subject Sample # 1	0.000	23:26
Breath Volume	2.046	23:26
Air Blank	0.000	23:27
Air Blank	0.000	23:29
Subject Sample # 2	0.000	23:29
Breath Volume	2.515	23:29
Air Blank	0.000	23:30
Control Test	0.081	23:30
Air Blank	0.000	23:31
Diagnostics Check	OK	23:31

Data Download Date/Time: 10/01/2019 00:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,TROY,M
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: KAISER,DAVID
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 09/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: JESUS

DATE/TIME OF TEST: 09/28/2019 22:25:00

Diagnostics Check	OK	22:17
Air Blank	0.000	22:17
Control Test	0.082	22:17
Air Blank	0.000	22:18
Subject Sample # 1	0.216	22:18
Breath Volume	1.164	22:18
Air Blank	0.000	22:19
Air Blank	0.000	22:21
Subject Sample # 2	0.210	22:23
Breath Volume	3.281	22:23
Air Blank	0.000	22:24
Control Test	0.076	22:24
Air Blank	0.000	22:24
Diagnostics Check	OK	22:25

Data Download Date/Time: 10/01/2019 00:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,TROY,M
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: GOODCHILD,JUSTIN
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 09/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: ANTOINE

DATE/TIME OF TEST: 09/30/2019 02:02:16

Diagnostics Check	OK	01:55
Air Blank	0.000	01:56
Control Test	0.081	01:56
Air Blank	0.000	01:56
Subject Sample # 1	0.173	01:57
Breath Volume	2.246	01:57
Air Blank	0.000	01:58
Air Blank	0.000	02:00
Subject Sample # 2	0.179	02:00
Breath Volume	2.089	02:00
Air Blank	0.000	02:01
Control Test	0.081	02:01
Air Blank	0.000	02:02
Diagnostics Check	OK	02:02

Data Download Date/Time: 10/01/2019 00:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,TROY,M
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: HOLLAND,TROY
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 09/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 19:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:59

SUBJECT'S LAST NAME: BRODERICK

DATE/TIME OF TEST: 09/30/2019 21:31:25

Diagnostics Check	OK	21:17	
Air Blank	0.000	21:18	
Control Test	0.081	21:18	
Air Blank	0.000	21:18	
Subject Sample	# 1 0.144	21:22	Volume Not Met
Breath Volume	0.351	21:22	
Air Blank	0.000	21:22	
Air Blank	0.000	21:24	
Subject Sample	# 2 0.178	21:26	
Breath Volume	2.289	21:26	
Air Blank	0.000	21:26	
Air Blank	0.000	21:28	
Subject Sample	# 3 0.162	21:29	
Breath Volume	1.410	21:29	
Air Blank	0.000	21:30	
Control Test	0.081	21:30	
Air Blank	0.000	21:31	
Diagnostics Check	OK	21:31	

Data Download Date/Time: 10/01/2019 00:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: MASSIP,LEONARDO
ARRESTING OFFICER AGENCY: Oviedo PD
LAST AGENCY INSPECTION: 09/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 09/30/2019 22:51:35

Diagnostics Check	OK	22:45
Air Blank	0.000	22:45
Control Test	0.081	22:45
Air Blank	0.000	22:46
Subject Sample # 1	0.143	22:46
Breath Volume	2.355	22:46
Air Blank	0.000	22:47
Air Blank	0.000	22:49
Subject Sample # 2	0.133	22:50
Breath Volume	2.320	22:50
Air Blank	0.000	22:50
Control Test	0.080	22:50
Air Blank	0.000	22:51
Diagnostics Check	OK	22:51

Data Download Date/Time: 10/01/2019 00:52