Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: TRAN

DATE/TIME OF TEST: 10/01/2020 04:08:17

Diagnostics Check		OK	04:00
Air Blank		0.000	04:01
Control Test		0.080	04:01
Air Blank		0.000	04:02
Subject Sample	# 1	0.175	04:03
Breath Volume		1.738	04:03
Air Blank		0.000	04:04
Air Blank		0.000	04:06
Subject Sample	# 2	0.176	04:06
Breath Volume		1.722	04:06
Air Blank		0.000	04:07
Control Test		0.079	04:07
Air Blank		0.000	04:08
Diagnostics Check		OK	04:08

Data Download Date/Time: 10/20/2020 09:53

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DELORENZO,D
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 10/01/2020 05:10:03

Diagnostics Check		OK	05:03
Air Blank		0.000	05:04
Control Test		0.079	05:04
Air Blank		0.000	05:05
Subject Sample	# 1	0.040	05:05
Breath Volume		1.406	05:05
Air Blank		0.000	05:06
Air Blank		0.000	05:07
Subject Sample	# 2	0.037	05:08
Breath Volume		1.625	05:08
Air Blank		0.000	05:09
Control Test		0.079	05:09
Air Blank		0.000	05:09
Diagnostics Check		OK	05:10

Data Download Date/Time: 10/20/2020 09:53

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RIVERA,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: RODNEY

DATE/TIME OF TEST: 10/02/2020 04:28:56

Diagnostics Check		OK	04:21
Air Blank		0.000	04:21
Control Test		0.081	04:22
Air Blank		0.000	04:22
Subject Sample	# 1	0.192	04:24
Breath Volume		2.167	04:24
Air Blank		0.000	04:24
Air Blank		0.000	04:26
Subject Sample	# 2	0.202	04:27
Breath Volume		2.832	04:27
Air Blank		0.000	04:28
Control Test		0.080	04:28
Air Blank		0.000	04:28
Diagnostics Check		OK	04:28

Data Download Date/Time: 10/20/2020 09:53

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RYAN,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: WATTS

DATE/TIME OF TEST: 10/03/2020 03:46:06

Diagnostics Check		OK	03:38
Air Blank		0.000	03:39
Control Test		0.080	03:39
Air Blank		0.000	03:39
Subject Sample	# 1	0.054	03:41
Breath Volume		1.441	03:41
Air Blank		0.000	03:41
Air Blank		0.000	03:43
Subject Sample	# 2	0.054	03:44
Breath Volume		1.429	03:44
Air Blank		0.000	03:45
Control Test		0.079	03:45
Air Blank		0.000	03:46
Diagnostics Check		OK	03:46

Data Download Date/Time: 10/20/2020 09:53

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: OCONNOR

DATE/TIME OF TEST: 10/04/2020 02:12:49

Diagnostics Check		OK	02:05
Air Blank		0.000	02:05
Control Test		0.080	02:06
Air Blank		0.000	02:06
Subject Sample	# 1	0.213	02:08
Breath Volume		3.886	02:08
Air Blank		0.000	02:08
Air Blank		0.000	02:10
Subject Sample	# 2	0.214	02:11
Breath Volume		3.882	02:11
Air Blank		0.000	02:11
Control Test		0.078	02:12
Air Blank		0.000	02:12
Diagnostics Check		OK	02:12

Data Download Date/Time: 10/20/2020 09:53

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: LOOMIS

DATE/TIME OF TEST: 10/05/2020 02:30:31

Diagnostics Check		OK	02:24
Air Blank		0.000	02:24
Control Test		0.080	02:25
Air Blank		0.000	02:25
Subject Sample	# 1	0.133	02:26
Breath Volume		2.500	02:26
Air Blank		0.000	02:26
Air Blank		0.000	02:28
Subject Sample	# 2	0.127	02:28
Breath Volume		1.488	02:28
Air Blank		0.000	02:29
Control Test		0.080	02:29
Air Blank		0.000	02:30
Diagnostics Check		OK	02:30

Data Download Date/Time: 10/20/2020 09:54

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HUTCHESON,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:26

SUBJECT'S LAST NAME: HESTER

DATE/TIME OF TEST: 10/07/2020 05:58:25

Diagnostics Check		OK	05:46	
Air Blank		0.000	05:47	
Control Test		0.080	05:47	
Air Blank		0.000	05:48	
Subject Sample	# 1	0.026	05:51	Volume Not Met
Breath Volume		0.132	05:51	
Air Blank		0.000	05:51	
Air Blank		0.000	05:53	
Subject Sample	# 2	0.023	05:56	Volume Not Met
Breath Volume		0.148	05:56	
Air Blank		0.000	05:57	
Control Test		0.079	05:57	
Air Blank		0.000	05:58	
Diagnostics Check		OK	05:58	

Data Download Date/Time: 10/20/2020 09:54

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HUTCHESON,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:54

SUBJECT'S LAST NAME: HESTER

DATE/TIME OF TEST: 10/07/2020 06:25:53

Diagnostics Check		OK	06:14	
Air Blank		0.000	06:14	
Control Test		0.080	06:15	
Air Blank		0.000	06:15	
Subject Sample	# 1	0.020	06:18	Volume Not Met
Breath Volume		0.324	06:18	
Air Blank		0.000	06:19	
Air Blank		0.000	06:21	
Subject Sample	# 2	0.018	06:24	Volume Not Met
Breath Volume		0.328	06:24	
Air Blank		0.000	06:24	
Control Test		0.080	06:25	
Air Blank		0.000	06:25	
Diagnostics Check		OK	06:25	

Data Download Date/Time: 10/20/2020 09:54

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PALFY,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: WIXON

DATE/TIME OF TEST: 10/08/2020 02:49:11

Diagnostics Check		OK	02:45	
Air Blank		0.000	02:45	
Control Test		0.079	02:46	
Air Blank		0.000	02:46	
Subject Sample	# 1		02:47	Subject Test Refused
Breath Volume		0.000	02:47	
Air Blank		0.000	02:48	
Control Test		0.080	02:48	
Air Blank		0.000	02:49	
Diagnostics Check		OK	02:49	

Data Download Date/Time: 10/20/2020 09:54

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: CRUZ-SANTIAGO

DATE/TIME OF TEST: 10/08/2020 22:33:16

Diagnostics Check		OK	22:26
Air Blank		0.000	22:26
Control Test		0.080	22:27
Air Blank		0.000	22:27
Subject Sample	# 1	0.178	22:28
Breath Volume		1.820	22:28
Air Blank		0.000	22:29
Air Blank		0.000	22:30
Subject Sample	# 2	0.158	22:31
Breath Volume		1.957	22:31
Air Blank		0.000	22:32
Control Test		0.079	22:32
Air Blank		0.000	22:33
Diagnostics Check		OK	22:33

Data Download Date/Time: 10/20/2020 09:54

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GROSS,R
ARRESTING OFFICER AGENCY: U of So FL Campus PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: ARENS

DATE/TIME OF TEST: 10/09/2020 02:59:28

Diagnostics Check		OK	02:52
Air Blank		0.000	02:53
Control Test		0.079	02:53
Air Blank		0.000	02:54
Subject Sample	# 1	0.195	02:54
Breath Volume		2.238	02:54
Air Blank		0.000	02:55
Air Blank		0.000	02:57
Subject Sample	# 2	0.190	02:57
Breath Volume		2.730	02:57
Air Blank		0.000	02:58
Control Test		0.078	02:58
Air Blank		0.000	02:59
Diagnostics Check		OK	02:59

Data Download Date/Time: 10/20/2020 09:54

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,BARBARA,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CRAVER,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:57

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 10/09/2020 18:25:07

Diagnostics Check		OK	18:18
Air Blank		0.000	18:18
Control Test		0.079	18:19
Air Blank		0.000	18:19
Subject Sample	# 1	0.146	18:20
Breath Volume		2.218	18:20
Air Blank		0.000	18:21
Air Blank		0.000	18:22
Subject Sample	# 2	0.148	18:23
Breath Volume		2.394	18:23
Air Blank		0.000	18:24
Control Test		0.079	18:24
Air Blank		0.000	18:25
Diagnostics Check		OK	18:25

Data Download Date/Time: 10/20/2020 09:54

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HARRIS,S
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: JONES JR

DATE/TIME OF TEST: 10/09/2020 22:09:14

Diagnostics Check		OK	22:03
Air Blank		0.000	22:03
Control Test		0.080	22:03
Air Blank		0.000	22:04
Subject Sample	# 1	0.000	22:04
Breath Volume		2.148	22:04
Air Blank		0.000	22:05
Air Blank		0.000	22:07
Subject Sample	# 2	0.000	22:07
Breath Volume		2.078	22:07
Air Blank		0.000	22:08
Control Test		0.080	22:08
Air Blank		0.000	22:09
Diagnostics Check		OK	22:09

Data Download Date/Time: 10/20/2020 09:54

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TIBURCIO,I
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: BARRAZA-BARAJA

DATE/TIME OF TEST: 10/10/2020 01:55:27

Diagnostics Check		OK	01:46
Air Blank		0.000	01:46
Control Test		0.080	01:47
Air Blank		0.000	01:47
Subject Sample	# 1	0.186	01:50
Breath Volume		1.578	01:50
Air Blank		0.000	01:51
Air Blank		0.000	01:53
Subject Sample	# 2	0.194	01:53
Breath Volume		1.359	01:53
Air Blank		0.000	01:54
Control Test		0.078	01:54
Air Blank		0.000	01:55
Diagnostics Check		OK	01:55

Data Download Date/Time: 10/20/2020 09:54

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HUTCHESON,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: EDWARDS III

DATE/TIME OF TEST: 10/10/2020 05:27:15

Diagnostics Check		OK	05:20
Air Blank		0.000	05:21
Control Test		0.078	05:21
Air Blank		0.000	05:22
Subject Sample	# 1	0.215	05:22
Breath Volume		2.902	05:22
Air Blank		0.000	05:23
Air Blank		0.000	05:25
Subject Sample	# 2	0.205	05:25
Breath Volume		1.570	05:25
Air Blank		0.000	05:26
Control Test		0.078	05:26
Air Blank		0.000	05:27
Diagnostics Check		OK	05:27

Data Download Date/Time: 10/20/2020 09:54

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: PRINCE

DATE/TIME OF TEST: 10/11/2020 00:56:33

Diagnostics Check		OK	00:50
Air Blank		0.000	00:50
Control Test		0.080	00:50
Air Blank		0.000	00:51
Subject Sample	# 1	0.096	00:52
Breath Volume		2.285	00:52
Air Blank		0.000	00:52
Air Blank		0.000	00:54
Subject Sample	# 2	0.094	00:54
Breath Volume		1.277	00:54
Air Blank		0.000	00:55
Control Test		0.080	00:55
Air Blank		0.000	00:56
Diagnostics Check		OK	00:56

Data Download Date/Time: 10/20/2020 09:54

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FISHER,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: HE

DATE/TIME OF TEST: 10/11/2020 03:56:32

Diagnostics Check		OK	03:46
Air Blank		0.000	03:46
Control Test		0.080	03:46
Air Blank		0.000	03:47
Subject Sample	# 1	0.092	03:50
Breath Volume		1.398	03:50
Air Blank		0.000	03:50
Air Blank		0.000	03:52
Subject Sample	# 2	0.096	03:54
Breath Volume		2.218	03:54
Air Blank		0.000	03:55
Control Test		0.078	03:55
Air Blank		0.000	03:56
Diagnostics Check		OK	03:56

Data Download Date/Time: 10/20/2020 09:55

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RIVERA,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: BENITEZ

DATE/TIME OF TEST: 10/11/2020 04:35:02

Diagnostics Check		OK	04:26
Air Blank		0.000	04:26
Control Test		0.079	04:27
Air Blank		0.000	04:27
Subject Sample	# 1	0.169	04:29
Breath Volume		1.398	04:29
Air Blank		0.000	04:30
Air Blank		0.000	04:31
Subject Sample	# 2	0.171	04:33
Breath Volume		1.867	04:33
Air Blank		0.000	04:34
Control Test		0.079	04:34
Air Blank		0.000	04:34
Diagnostics Check		OK	04:35

Data Download Date/Time: 10/20/2020 09:55

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ARCHAMBEAU,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: SNYDER

DATE/TIME OF TEST: 10/13/2020 04:39:47

Diagnostics Check		OK	04:28	
Air Blank		0.000	04:28	
Control Test		0.080	04:28	
Air Blank		0.000	04:29	
Subject Sample	# 1	0.138	04:32	Volume Not Met
Breath Volume		0.527	04:32	
Air Blank		0.000	04:33	
Air Blank		0.000	04:34	
Subject Sample	# 2	0.138	04:38	Volume Not Met
Breath Volume		0.261	04:38	
Air Blank		0.000	04:38	
Control Test		0.079	04:39	
Air Blank		0.000	04:39	
Diagnostics Check		OK	04:39	

Data Download Date/Time: 10/20/2020 09:55

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ARCHAMBEAU,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: SNYDER

DATE/TIME OF TEST: 10/13/2020 04:51:10

Diagnostics Check		OK	04:41
Air Blank		0.000	04:42
Control Test		0.079	04:42
Air Blank		0.000	04:43
Subject Sample	# 1	0.146	04:45
Breath Volume		1.433	04:45
Air Blank		0.000	04:46
Air Blank		0.000	04:48
Subject Sample	# 2	0.159	04:49
Breath Volume		1.457	04:49
Air Blank		0.000	04:50
Control Test		0.078	04:50
Air Blank		0.000	04:51
Diagnostics Check		OK	04:51

Data Download Date/Time: 10/20/2020 09:55

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: COSTA,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 10/14/2020 01:53:09

Diagnostics Check		OK	01:46
Air Blank		0.000	01:47
Control Test		0.079	01:47
Air Blank		0.000	01:48
Subject Sample	# 1	0.076	01:48
Breath Volume		3.785	01:48
Air Blank		0.000	01:49
Air Blank		0.000	01:51
Subject Sample	# 2	0.074	01:51
Breath Volume		4.062	01:51
Air Blank		0.000	01:52
Control Test		0.079	01:52
Air Blank		0.000	01:53
Diagnostics Check		OK	01:53

Data Download Date/Time: 10/20/2020 09:55

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ATKINS,A
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: MCINTYRE

DATE/TIME OF TEST: 10/14/2020 04:41:54

Diagnostics Check		OK	04:34
Air Blank		0.000	04:34
Control Test		0.080	04:35
Air Blank		0.000	04:35
Subject Sample	# 1	0.167	04:37
Breath Volume		1.976	04:37
Air Blank		0.000	04:38
Air Blank		0.000	04:39
Subject Sample	# 2	0.173	04:40
Breath Volume		1.316	04:40
Air Blank		0.000	04:40
Control Test		0.080	04:41
Air Blank		0.000	04:41
Diagnostics Check		OK	04:41

Data Download Date/Time: 10/20/2020 09:55

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: KUCHTA

DATE/TIME OF TEST: 10/15/2020 02:32:12

Diagnostics Check		OK	02:25
Air Blank		0.000	02:25
Control Test		0.080	02:26
Air Blank		0.000	02:26
Subject Sample	# 1	0.122	02:27
Breath Volume		2.363	02:27
Air Blank		0.000	02:27
Air Blank		0.000	02:29
Subject Sample	# 2	0.119	02:30
Breath Volume		2.500	02:30
Air Blank		0.000	02:31
Control Test		0.080	02:31
Air Blank		0.000	02:32
Diagnostics Check		OK	02:32

Data Download Date/Time: 10/20/2020 09:55

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: REID,W
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:59

SUBJECT'S LAST NAME: LINARES

DATE/TIME OF TEST: 10/15/2020 05:29:46

Diagnostics Check		OK	05:21
Air Blank		0.000	05:21
Control Test		0.080	05:22
Air Blank		0.000	05:22
Subject Sample	# 1	0.197	05:24
Breath Volume		1.890	05:24
Air Blank		0.000	05:25
Air Blank		0.000	05:27
Subject Sample	# 2	0.199	05:28
Breath Volume		2.796	05:28
Air Blank		0.000	05:28
Control Test		0.079	05:29
Air Blank		0.000	05:29
Diagnostics Check		OK	05:29

Data Download Date/Time: 10/20/2020 09:55

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TIBURCIO,I
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 10/16/2020 04:25:00

Diagnostics Check		OK	04:18
Air Blank		0.000	04:18
Control Test		0.080	04:19
Air Blank		0.000	04:19
Subject Sample	# 1	0.148	04:20
Breath Volume		4.085	04:20
Air Blank		0.000	04:21
Air Blank		0.000	04:23
Subject Sample	# 2	0.143	04:23
Breath Volume		3.953	04:23
Air Blank		0.000	04:24
Control Test		0.080	04:24
Air Blank		0.000	04:24
Diagnostics Check		OK	04:25

Data Download Date/Time: 10/20/2020 09:55

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARBER,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:32

SUBJECT'S LAST NAME: FERGUSON

DATE/TIME OF TEST: 10/16/2020 20:04:11

Diagnostics Check		OK	19:52	
Air Blank		0.000	19:52	
Control Test		0.079	19:53	
Air Blank		0.000	19:53	
Subject Sample	# 1	0.154	19:56	Volume Not Met
Breath Volume		0.449	19:56	
Air Blank		0.000	19:57	
Air Blank		0.000	19:59	
Subject Sample	# 2	0.150	20:02	Volume Not Met
Breath Volume		0.140	20:02	
Air Blank		0.000	20:03	
Control Test		0.078	20:03	
Air Blank		0.000	20:04	
Diagnostics Check		OK	20:04	

Data Download Date/Time: 10/20/2020 09:55

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARBER,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:32

SUBJECT'S LAST NAME: FERGUSON

DATE/TIME OF TEST: 10/16/2020 20:16:14

Diagnostics Check		OK	20:07	
Air Blank		0.000	20:08	
Control Test		0.078	20:08	
Air Blank		0.000	20:09	
Subject Sample	# 1	0.134	20:12	Volume Not Met
Breath Volume		0.078	20:12	
Air Blank		0.000	20:12	
Air Blank		0.000	20:14	
Subject Sample	# 2		20:14	Subject Test Refused
Breath Volume		0.000	20:14	
Air Blank		0.000	20:15	
Control Test		0.077	20:15	
Air Blank		0.000	20:16	
Diagnostics Check		OK	20:16	

Data Download Date/Time: 10/20/2020 09:56

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: POOLER

DATE/TIME OF TEST: 10/17/2020 03:17:13

Diagnostics Check		OK	03:14	
Air Blank		0.000	03:14	
Control Test		0.079	03:14	
Air Blank		0.000	03:15	
Subject Sample	# 1		03:15	Subject Test Refused
Breath Volume		0.000	03:15	
Air Blank		0.000	03:16	
Control Test		0.080	03:16	
Air Blank		0.000	03:17	
Diagnostics Check		OK	03:17	

Data Download Date/Time: 10/20/2020 09:56

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: GALLAGLY

DATE/TIME OF TEST: 10/17/2020 04:00:44

Diagnostics Check		OK	03:54
Air Blank		0.000	03:54
Control Test		0.079	03:54
Air Blank		0.000	03:55
Subject Sample	# 1	0.163	03:56
Breath Volume		2.527	03:56
Air Blank		0.000	03:56
Air Blank		0.000	03:58
Subject Sample	# 2	0.166	03:59
Breath Volume		4.039	03:59
Air Blank		0.000	03:59
Control Test		0.079	04:00
Air Blank		0.000	04:00
Diagnostics Check		OK	04:00

Data Download Date/Time: 10/20/2020 09:56

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:58

SUBJECT'S LAST NAME: DOERN

DATE/TIME OF TEST: 10/17/2020 21:27:09

Diagnostics Check		OK	21:19
Air Blank		0.000	21:20
Control Test		0.080	21:20
Air Blank		0.000	21:20
Subject Sample	# 1	0.188	21:21
Breath Volume		1.203	21:21
Air Blank		0.000	21:22
Air Blank		0.000	21:23
Subject Sample	# 2	0.189	21:25
Breath Volume		1.359	21:25
Air Blank		0.000	21:26
Control Test		0.080	21:26
Air Blank		0.000	21:27
Diagnostics Check		OK	21:27

Data Download Date/Time: 10/20/2020 09:56

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: CAREY

DATE/TIME OF TEST: 10/18/2020 02:01:34

Diagnostics Check		OK	01:54
Air Blank		0.000	01:55
Control Test		0.080	01:55
Air Blank		0.000	01:55
Subject Sample	# 1	0.206	01:56
Breath Volume		2.980	01:56
Air Blank		0.000	01:57
Air Blank		0.000	01:59
Subject Sample	# 2	0.210	01:59
Breath Volume		2.500	01:59
Air Blank		0.000	02:00
Control Test		0.079	02:00
Air Blank		0.000	02:01
Diagnostics Check		OK	02:01

Data Download Date/Time: 10/20/2020 09:56

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PORTMAN,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 10/18/2020 02:14:53

Diagnostics Check		OK	02:08
Air Blank		0.000	02:08
Control Test		0.080	02:09
Air Blank		0.000	02:09
Subject Sample	# 1	0.157	02:10
Breath Volume		2.597	02:10
Air Blank		0.000	02:11
Air Blank		0.000	02:12
Subject Sample	# 2	0.160	02:13
Breath Volume		2.500	02:13
Air Blank		0.000	02:13
Control Test		0.079	02:14
Air Blank		0.000	02:14
Diagnostics Check		OK	02:14

Data Download Date/Time: 10/20/2020 09:56

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: TREVINO

DATE/TIME OF TEST: 10/18/2020 04:06:42

Diagnostics Check		OK	04:00
Air Blank		0.000	04:00
Control Test		0.080	04:01
Air Blank		0.000	04:01
Subject Sample	# 1	0.040	04:02
Breath Volume		2.515	04:02
Air Blank		0.000	04:02
Air Blank		0.000	04:04
Subject Sample	# 2	0.039	04:05
Breath Volume		2.082	04:05
Air Blank		0.000	04:05
Control Test		0.079	04:06
Air Blank		0.000	04:06
Diagnostics Check		OK	04:06

Data Download Date/Time: 10/20/2020 09:56

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 10/18/2020 05:27:50

Diagnostics Check		OK	05:20
Air Blank		0.000	05:20
Control Test		0.080	05:20
Air Blank		0.000	05:21
Subject Sample	# 1	0.174	05:23
Breath Volume		2.480	05:23
Air Blank		0.000	05:23
Air Blank		0.000	05:25
Subject Sample	# 2	0.173	05:26
Breath Volume		1.996	05:26
Air Blank		0.000	05:26
Control Test		0.079	05:27
Air Blank		0.000	05:27
Diagnostics Check		OK	05:27

Data Download Date/Time: 10/20/2020 09:56

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKIPPER,ROGER,G
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RUIZ,J 4481
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 06:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:52

SUBJECT'S LAST NAME: MENDEZ HERNANDEZ

DATE/TIME OF TEST: 10/19/2020 07:22:16

Diagnostics Check		OK	07:14
Air Blank		0.000	07:15
Control Test		0.079	07:15
Air Blank		0.000	07:15
Subject Sample	# 1	0.200	07:17
Breath Volume		2.765	07:17
Air Blank		0.000	07:18
Air Blank		0.000	07:20
Subject Sample	# 2	0.202	07:20
Breath Volume		1.554	07:20
Air Blank		0.000	07:21
Control Test		0.078	07:21
Air Blank		0.000	07:22
Diagnostics Check		OK	07:22

Data Download Date/Time: 10/20/2020 09:56

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,ANGELA,V
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PAYNTER,HUNTER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:09

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 10/19/2020 07:36:23

Diagnostics Check		OK	07:29
Air Blank		0.000	07:29
Control Test		0.080	07:30
Air Blank		0.000	07:30
Subject Sample	# 1	0.113	07:31
Breath Volume		1.371	07:31
Air Blank		0.000	07:32
Air Blank		0.000	07:33
Subject Sample	# 2	0.112	07:34
Breath Volume		1.597	07:34
Air Blank		0.000	07:35
Control Test		0.079	07:35
Air Blank		0.000	07:36
Diagnostics Check		OK	07:36

Data Download Date/Time: 10/20/2020 09:56

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NICOLAS,MITA,
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: GADDIE,KYLE
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 09/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: GARLAND

DATE/TIME OF TEST: 10/02/2020 23:48:46

Diagnostics Check		OK	23:39
Air Blank		0.000	23:40
Control Test		0.081	23:40
Air Blank		0.000	23:40
Subject Sample	# 1	0.268	23:41
Breath Volume		2.656	23:41
Air Blank		0.000	23:41
Air Blank		0.000	23:43
Subject Sample	# 2	0.244	23:44
Breath Volume		1.222	23:44
Air Blank		0.000	23:45
Air Blank		0.000	23:46
Subject Sample	# 3	0.277	23:47
Breath Volume		1.710	23:47
Air Blank		0.000	23:47
Control Test		0.079	23:48
Air Blank		0.000	23:48
Diagnostics Check		OK	23:48

Data Download Date/Time: 11/05/2020 13:05

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NICOLAS,MITA,
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: SCHERR,EVAN
ARRESTING OFFICER AGENCY: Flagler Beach PD
LAST AGENCY INSPECTION: 09/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: MORENO LOPEZ

DATE/TIME OF TEST: 10/03/2020 01:06:50

Diagnostics Check		OK	00:55	
Air Blank		0.000	00:55	
Control Test		0.081	00:55	
Air Blank		0.000	00:56	
Subject Sample	# 1	0.191	00:59	Slope Not Level
Breath Volume		1.394	00:59	
Air Blank		0.000	01:00	
Air Blank		0.000	01:02	
Subject Sample	# 2	0.176	01:05	Volume Not Met
Breath Volume		0.457	01:05	
Air Blank		0.000	01:05	
Control Test		0.079	01:06	
Air Blank		0.000	01:06	
Diagnostics Check		OK	01:06	

Data Download Date/Time: 11/05/2020 13:05

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NICOLAS,MITA,
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: SCHERR,EVAN
ARRESTING OFFICER AGENCY: Flagler Beach PD
LAST AGENCY INSPECTION: 09/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: MORENO LOPEZ

DATE/TIME OF TEST: 10/03/2020 01:16:04

Diagnostics Check		OK	01:09
Air Blank		0.000	01:10
Control Test		0.079	01:10
Air Blank		0.000	01:10
Subject Sample	# 1	0.202	01:11
Breath Volume		1.156	01:11
Air Blank		0.000	01:12
Air Blank		0.000	01:14
Subject Sample	# 2	0.208	01:14
Breath Volume		1.269	01:14
Air Blank		0.000	01:15
Control Test		0.079	01:15
Air Blank		0.000	01:15
Diagnostics Check		OK	01:16

Data Download Date/Time: 11/05/2020 13:05

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESTEP,JEFFREY,D
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: GUERRERO,M
ARRESTING OFFICER AGENCY: Flagler Beach PD
LAST AGENCY INSPECTION: 09/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: PARDINY

DATE/TIME OF TEST: 10/03/2020 03:43:08

Diagnostics Check		OK	03:36
Air Blank		0.000	03:36
Control Test		0.080	03:36
Air Blank		0.000	03:37
Subject Sample	# 1	0.169	03:38
Breath Volume		3.507	03:38
Air Blank		0.000	03:39
Air Blank		0.000	03:40
Subject Sample	# 2	0.171	03:41
Breath Volume		2.863	03:41
Air Blank		0.000	03:42
Control Test		0.077	03:42
Air Blank		0.000	03:43
Diagnostics Check		OK	03:43

Data Download Date/Time: 11/05/2020 13:05

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,CARL,L
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: PARKER,CARL
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 09/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: CUNNINGHAM

DATE/TIME OF TEST: 10/03/2020 23:18:57

Diagnostics Check		OK	23:16	
Air Blank		0.000	23:16	
Control Test		0.080	23:16	
Air Blank		0.000	23:17	
Subject Sample	# 1		23:17	Subject Test Refused
Breath Volume		0.000	23:17	
Air Blank		0.000	23:18	
Control Test		0.081	23:18	
Air Blank		0.000	23:18	
Diagnostics Check		OK	23:18	

Data Download Date/Time: 11/05/2020 13:05

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NICOLAS,MITA,
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: SCHERR,EVAN
ARRESTING OFFICER AGENCY: Flagler Beach PD
LAST AGENCY INSPECTION: 09/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: MYERS

DATE/TIME OF TEST: 10/03/2020 23:29:29

Diagnostics Check		OK	23:23
Air Blank		0.000	23:23
Control Test		0.081	23:24
Air Blank		0.000	23:24
Subject Sample	# 1	0.106	23:24
Breath Volume		1.871	23:24
Air Blank		0.000	23:25
Air Blank		0.000	23:27
Subject Sample	# 2	0.105	23:27
Breath Volume		2.128	23:27
Air Blank		0.000	23:28
Control Test		0.080	23:28
Air Blank		0.000	23:29
Diagnostics Check		OK	23:29

Data Download Date/Time: 11/05/2020 13:05

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,CARL,L
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: PARKER,CARL
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 09/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: DASILVA

DATE/TIME OF TEST: 10/05/2020 01:14:46

Diagnostics Check		OK	01:08
Air Blank		0.000	01:08
Control Test		0.081	01:09
Air Blank		0.000	01:09
Subject Sample	# 1	0.172	01:10
Breath Volume		3.132	01:10
Air Blank		0.000	01:10
Air Blank		0.000	01:12
Subject Sample	# 2	0.167	01:13
Breath Volume		2.128	01:13
Air Blank		0.000	01:13
Control Test		0.080	01:14
Air Blank		0.000	01:14
Diagnostics Check		OK	01:14

Data Download Date/Time: 11/05/2020 13:05

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,CARL,L
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: PARKER,CARL
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 09/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: HADEN

DATE/TIME OF TEST: 10/09/2020 00:29:22

Diagnostics Check		OK	00:26	
Air Blank		0.000	00:26	
Control Test		0.081	00:27	
Air Blank		0.000	00:27	
Subject Sample	# 1		00:27	Subject Test Refused
Breath Volume		0.000	00:27	
Air Blank		0.000	00:28	
Control Test		0.080	00:28	
Air Blank		0.000	00:29	
Diagnostics Check		OK	00:29	

Data Download Date/Time: 11/05/2020 13:05

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,CARL,L
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: PARKER,CARL
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 09/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: GOLDSMITH HARVEY

DATE/TIME OF TEST: 10/17/2020 00:53:22

Diagnostics Check		OK	00:47
Air Blank		0.000	00:47
Control Test		0.081	00:47
Air Blank		0.000	00:48
Subject Sample	# 1	0.216	00:48
Breath Volume		1.519	00:48
Air Blank		0.000	00:49
Air Blank		0.000	00:51
Subject Sample	# 2	0.219	00:51
Breath Volume		1.816	00:51
Air Blank		0.000	00:52
Control Test		0.080	00:52
Air Blank		0.000	00:53
Diagnostics Check		OK	00:53

Data Download Date/Time: 11/05/2020 13:05

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESTEP,JEFFREY,D
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: PREVATT,JENNIFER
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 09/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:29

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 10/17/2020 19:00:31

Diagnostics Check		OK	18:49
Air Blank		0.000	18:49
Control Test		0.081	18:49
Air Blank		0.000	18:50
Subject Sample	# 1	0.148	18:51
Breath Volume		1.386	18:51
Air Blank		0.000	18:52
Air Blank		0.000	18:53
Subject Sample	# 2	0.123	18:54
Breath Volume		1.222	18:54
Air Blank		0.000	18:54
Air Blank		0.000	18:56
Subject Sample	# 3	0.129	18:58
Breath Volume		1.449	18:58
Air Blank		0.000	18:59
Control Test		0.079	18:59
Air Blank		0.000	19:00
Diagnostics Check		OK	19:00

Data Download Date/Time: 11/05/2020 13:05

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NICOLAS,MITA,
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: RAIHALA,J
ARRESTING OFFICER AGENCY: Bunnell PD
LAST AGENCY INSPECTION: 09/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: KEENE

DATE/TIME OF TEST: 10/18/2020 02:56:49

Diagnostics Check		OK	02:50
Air Blank		0.000	02:51
Control Test		0.081	02:51
Air Blank		0.000	02:51
Subject Sample	# 1	0.065	02:52
Breath Volume		1.226	02:52
Air Blank		0.000	02:52
Air Blank		0.000	02:54
Subject Sample	# 2	0.067	02:55
Breath Volume		1.191	02:55
Air Blank		0.000	02:55
Control Test		0.081	02:56
Air Blank		0.000	02:56
Diagnostics Check		OK	02:56

Data Download Date/Time: 11/05/2020 13:05

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRIEMER,AMANDA,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: MORGANTE,T
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 09/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: LAZZARO

DATE/TIME OF TEST: 10/24/2020 04:16:06

Diagnostics Check		OK	04:11	
Air Blank		0.000	04:12	
Control Test		0.081	04:12	
Air Blank		0.000	04:13	
Subject Sample	# 1	0.150	04:14	
Breath Volume		1.605	04:14	
Air Blank			04:15	Purge Fail
Air Blank			04:16	Purge Fail

Data Download Date/Time: 11/05/2020 13:05

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRIEMER,AMANDA,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: MORGANTE,T
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 09/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: LAZZARO

DATE/TIME OF TEST: 10/24/2020 04:35:32

Diagnostics Check		OK	04:29
Air Blank		0.000	04:29
Control Test		0.079	04:29
Air Blank		0.000	04:30
Subject Sample	# 1	0.136	04:31
Breath Volume		1.562	04:31
Air Blank		0.000	04:31
Air Blank		0.000	04:33
Subject Sample	# 2	0.150	04:33
Breath Volume		1.550	04:33
Air Blank		0.000	04:34
Control Test		0.078	04:34
Air Blank		0.000	04:35
Diagnostics Check		OK	04:35

Data Download Date/Time: 11/05/2020 13:05

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAUN,R,A
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: WICHMAN,M
ARRESTING OFFICER AGENCY: Bunnell PD
LAST AGENCY INSPECTION: 09/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 15:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:34

SUBJECT'S LAST NAME: RIGNEY

DATE/TIME OF TEST: 10/24/2020 17:01:36

Diagnostics Check		OK	16:55
Air Blank		0.000	16:55
Control Test		0.080	16:56
Air Blank		0.000	16:56
Subject Sample	# 1	0.290	16:56
Breath Volume		3.101	16:56
Air Blank		0.000	16:57
Air Blank		0.000	16:59
Subject Sample	# 2	0.276	16:59
Breath Volume		2.160	16:59
Air Blank		0.000	17:00
Control Test		0.078	17:00
Air Blank		0.000	17:01
Diagnostics Check		OK	17:01

Data Download Date/Time: 11/05/2020 13:05

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRIEMER,AMANDA,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: CLAY,A
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 09/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: DEROO

DATE/TIME OF TEST: 10/24/2020 23:43:25

Diagnostics Check		OK	23:36
Air Blank		0.000	23:37
Control Test		0.081	23:37
Air Blank		0.000	23:38
Subject Sample	# 1	0.074	23:38
Breath Volume		1.355	23:38
Air Blank		0.000	23:39
Air Blank		0.000	23:41
Subject Sample	# 2	0.076	23:41
Breath Volume		1.714	23:41
Air Blank		0.000	23:42
Control Test		0.080	23:42
Air Blank		0.000	23:43
Diagnostics Check		OK	23:43

Data Download Date/Time: 11/05/2020 13:05

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,CARL,L
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: PARKER,CARL
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 09/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: PLUMMER

DATE/TIME OF TEST: 10/27/2020 04:03:33

Diagnostics Check		OK	03:57
Air Blank		0.000	03:57
Control Test		0.080	03:57
Air Blank		0.000	03:58
Subject Sample	# 1	0.135	03:58
Breath Volume		2.578	03:58
Air Blank		0.000	03:59
Air Blank		0.000	04:01
Subject Sample	# 2	0.130	04:01
Breath Volume		1.878	04:01
Air Blank		0.000	04:02
Control Test		0.080	04:02
Air Blank		0.000	04:03
Diagnostics Check		OK	04:03

Data Download Date/Time: 11/05/2020 13:05

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRIEMER,AMANDA,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: BIFANO,A
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 09/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: SARINANA

DATE/TIME OF TEST: 10/28/2020 22:10:52

Diagnostics Check		OK	22:05	
Air Blank		0.000	22:05	
Control Test		0.079	22:06	
Air Blank		0.000	22:06	
Subject Sample	# 1		22:09	Subject Test Refused
Breath Volume		0.000	22:09	
Air Blank		0.000	22:09	
Control Test		0.079	22:10	
Air Blank		0.000	22:10	
Diagnostics Check		OK	22:10	

Data Download Date/Time: 11/05/2020 13:05

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006628

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,VICTOR,J
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: RIVERA,VICTOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: RODARTE

DATE/TIME OF TEST: 10/31/2020 03:02:11

Diagnostics Check	OK	03:01	
Air Blank		03:01	Ambient Fail
Air Blank	0.000	03:02	

Data Download Date/Time: 11/09/2020 12:34

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006628

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,VICTOR,J
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: RIVERA,VICTOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: RODARTE

DATE/TIME OF TEST: 10/31/2020 03:10:09

Diagnostics Check		OK	03:03
Air Blank		0.000	03:04
Control Test		0.079	03:04
Air Blank		0.000	03:05
Subject Sample	# 1	0.155	03:05
Breath Volume		3.046	03:05
Air Blank		0.000	03:06
Air Blank		0.000	03:08
Subject Sample	# 2	0.154	03:08
Breath Volume		3.140	03:08
Air Blank		0.000	03:09
Control Test		0.080	03:09
Air Blank		0.000	03:10
Diagnostics Check		OK	03:10

Data Download Date/Time: 11/09/2020 12:34

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006629

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAWKINS,DEBORAH,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: MOONEY,RYAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1791908A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: GRANT

DATE/TIME OF TEST: 10/17/2020 05:09:27

Diagnostics Check		OK	05:01
Air Blank		0.000	05:01
Control Test		0.077	05:01
Air Blank		0.000	05:02
Subject Sample	# 1	0.203	05:04
Breath Volume		2.062	05:04
Air Blank		0.000	05:05
Air Blank		0.000	05:06
Subject Sample	# 2	0.200	05:07
Breath Volume		1.820	05:07
Air Blank		0.000	05:08
Control Test		0.075	05:08
Air Blank		0.000	05:09
Diagnostics Check		OK	05:09

Data Download Date/Time: 11/09/2020 12:27

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006629

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARPENTER,BLAKE,E
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: CARPENTER,BLAKE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1791908A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: SANCHEZ-LONGO

DATE/TIME OF TEST: 10/30/2020 00:20:11

Diagnostics Check		OK	00:04	
Air Blank		0.000	00:04	
Control Test		0.077	00:05	
Air Blank		0.000	00:05	
Subject Sample	# 1	0.151	00:08	Volume Not Met
Breath Volume		0.355	00:08	
Air Blank		0.000	00:09	
Air Blank		0.000	00:11	
Subject Sample	# 2	0.180	00:13	
Breath Volume		1.929	00:13	
Air Blank		0.000	00:14	
Air Blank		0.000	00:15	
Subject Sample	# 3	0.172	00:18	
Breath Volume		1.718	00:18	
Air Blank		0.000	00:19	
Control Test		0.075	00:19	
Air Blank		0.000	00:20	
Diagnostics Check		OK	00:20	

Data Download Date/Time: 11/09/2020 12:27

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRERO,FRAICOR,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TERRERO,FRAICOR
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 09/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: OATES

DATE/TIME OF TEST: 10/04/2020 23:07:35

Diagnostics Check		OK	23:01
Air Blank		0.000	23:02
Control Test		0.078	23:02
Air Blank		0.000	23:03
Subject Sample	# 1	0.000	23:03
Breath Volume		1.890	23:03
Air Blank		0.000	23:03
Air Blank		0.000	23:05
Subject Sample	# 2	0.000	23:06
Breath Volume		1.851	23:06
Air Blank		0.000	23:06
Control Test		0.078	23:06
Air Blank		0.000	23:07
Diagnostics Check		OK	23:07

Data Download Date/Time: 10/19/2020 10:45

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGBY,CHARLES,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BIGBY,CHARLES
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 09/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: SPENCER

DATE/TIME OF TEST: 10/13/2020 23:05:48

Diagnostics Check	OK	23:04	
Air Blank		23:05	RFI Detect
Air Blank	0.000	23:05	

Data Download Date/Time: 10/19/2020 10:45

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGBY,CHARLES,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BIGBY,CHARLES
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 09/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: SPENCER

DATE/TIME OF TEST: 10/13/2020 23:10:40

Diagnostics Check		OK	23:07	
Air Blank		0.000	23:08	
Control Test		0.079	23:08	
Air Blank		0.000	23:09	
Subject Sample	# 1		23:09	Subject Test Refused
Breath Volume		0.000	23:09	
Air Blank		0.000	23:09	
Control Test		0.079	23:10	
Air Blank		0.000	23:10	
Diagnostics Check		OK	23:10	

Data Download Date/Time: 10/19/2020 10:46

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORD,JOSHUA,N
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LORD,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 10/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: DUI
ARREST TIME: 16:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:10

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 10/20/2020 17:45:26

Diagnostics Check		OK	17:38
Air Blank		0.000	17:38
Control Test		0.078	17:39
Air Blank		0.000	17:39
Subject Sample	# 1	0.227	17:40
Breath Volume		2.050	17:40
Air Blank		0.000	17:41
Air Blank		0.000	17:43
Subject Sample	# 2	0.218	17:43
Breath Volume		2.410	17:43
Air Blank		0.000	17:44
Control Test		0.078	17:44
Air Blank		0.000	17:45
Diagnostics Check		OK	17:45

Data Download Date/Time: 11/24/2020 08:41

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EMRICH,IAN,W
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: EMRICH,IAN
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 10/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: DUI
ARREST TIME: 21:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: SHINHOLSTER

DATE/TIME OF TEST: 10/21/2020 22:06:37

Diagnostics Check	OK	22:05	
Air Blank		22:06	RFI Detect
Air Blank	0.000	22:06	

Data Download Date/Time: 11/24/2020 08:41

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAEZ,HAROLD,
BREATH TEST OPERATOR AGENCY: Palmetto PD
ARRESTING OFFICER: BAEZ,HAROLD
ARRESTING OFFICER AGENCY: Palmetto PD
LAST AGENCY INSPECTION: 10/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: DUI
ARREST TIME: 07:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:30

SUBJECT'S LAST NAME: ESCALANTE VELASQUEZ

DATE/TIME OF TEST: 10/24/2020 07:56:12

Diagnostics Check		OK	07:51	
Air Blank		0.000	07:51	
Control Test		0.079	07:52	
Air Blank		0.000	07:52	
Subject Sample	# 1		07:55	Sequence Aborted
Breath Volume		0.000	07:55	
Air Blank		0.000	07:56	

Data Download Date/Time: 11/24/2020 08:42

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAEZ,HAROLD,
BREATH TEST OPERATOR AGENCY: Palmetto PD
ARRESTING OFFICER: MOORE,ADAM
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 10/19/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: DUI
ARREST TIME: 09:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:50

SUBJECT'S LAST NAME: PETERS

DATE/TIME OF TEST: 10/24/2020 10:19:01

Diagnostics Check		OK	10:10
Air Blank		0.000	10:11
Control Test		0.079	10:11
Air Blank		0.000	10:11
Subject Sample	# 1	0.066	10:14
Breath Volume		1.417	10:14
Air Blank		0.000	10:14
Air Blank		0.000	10:16
Subject Sample	# 2	0.067	10:17
Breath Volume		1.148	10:17
Air Blank		0.000	10:18
Control Test		0.077	10:18
Air Blank		0.000	10:18
Diagnostics Check		OK	10:19

Data Download Date/Time: 11/24/2020 08:42

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHANKS,DAVID,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: FLORES,DAVID
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 10/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: SHANKS

DATE/TIME OF TEST: 10/26/2020 02:26:56

Diagnostics Check		OK	02:20
Air Blank		0.000	02:21
Control Test		0.080	02:21
Air Blank		0.000	02:21
Subject Sample	# 1	0.055	02:22
Breath Volume		3.101	02:22
Air Blank		0.000	02:23
Air Blank		0.000	02:24
Subject Sample	# 2	0.053	02:25
Breath Volume		3.312	02:25
Air Blank		0.000	02:25
Control Test		0.077	02:26
Air Blank		0.000	02:26
Diagnostics Check		OK	02:26

Data Download Date/Time: 11/24/2020 08:42

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006633

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAVALLE,D,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: LAVALLE,D
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: GORDON

DATE/TIME OF TEST: 10/30/2020 23:19:53

Diagnostics Check		OK	23:16	
Air Blank		0.000	23:17	
Control Test		0.080	23:17	
Air Blank		0.000	23:18	
Subject Sample	# 1		23:18	Subject Test Refused
Breath Volume		0.000	23:18	
Air Blank		0.000	23:18	
Control Test		0.081	23:19	
Air Blank		0.000	23:19	
Diagnostics Check		OK	23:19	

Data Download Date/Time: 11/25/2020 11:54

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,CODY,R
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: KOPE,MICHAEL
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 09/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: KLEIN

DATE/TIME OF TEST: 10/02/2020 23:16:56

Diagnostics Check		OK	23:09
Air Blank		0.000	23:10
Control Test		0.079	23:10
Air Blank		0.000	23:11
Subject Sample	# 1	0.302	23:12
Breath Volume		1.968	23:12
Air Blank		0.000	23:13
Air Blank		0.000	23:14
Subject Sample	# 2	0.288	23:15
Breath Volume		1.734	23:15
Air Blank		0.000	23:16
Control Test		0.075	23:16
Air Blank		0.000	23:16
Diagnostics Check		OK	23:16

Data Download Date/Time: 10/20/2020 08:54

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,CODY,R
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BARRETT,CODY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 09/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 10/04/2020 00:16:39

Diagnostics Check		OK	00:09
Air Blank		0.000	00:10
Control Test		0.079	00:10
Air Blank		0.000	00:11
Subject Sample	# 1	0.268	00:11
Breath Volume		2.683	00:11
Air Blank		0.000	00:12
Air Blank		0.000	00:14
Subject Sample	# 2	0.270	00:14
Breath Volume		3.265	00:14
Air Blank		0.000	00:15
Control Test		0.077	00:16
Air Blank		0.000	00:16
Diagnostics Check		OK	00:16

Data Download Date/Time: 10/20/2020 08:54

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,CODY,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BARRETT,CODY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 09/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 16:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: KINGMAN

DATE/TIME OF TEST: 10/04/2020 17:47:04

Diagnostics Check		OK	17:38
Air Blank		0.000	17:38
Control Test		0.079	17:38
Air Blank		0.000	17:39
Subject Sample	# 1	0.220	17:42
Breath Volume		2.546	17:42
Air Blank		0.000	17:43
Air Blank		0.000	17:44
Subject Sample	# 2	0.222	17:45
Breath Volume		1.812	17:45
Air Blank		0.000	17:46
Control Test		0.076	17:46
Air Blank		0.000	17:46
Diagnostics Check		OK	17:47

Data Download Date/Time: 10/20/2020 08:54

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTANA,ELIER,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SANTANA,ELIER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 09/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 11:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:24

SUBJECT'S LAST NAME: TISONE

DATE/TIME OF TEST: 10/07/2020 12:58:48

Diagnostics Check		OK	12:46
Air Blank		0.000	12:47
Control Test		0.079	12:47
Air Blank		0.000	12:48
Subject Sample	# 1	0.254	12:49
Breath Volume		1.683	12:49
Air Blank		0.000	12:50
Air Blank		0.000	12:52
Subject Sample	# 2	0.229	12:53
Breath Volume		1.347	12:53
Air Blank		0.000	12:54
Air Blank		0.000	12:56
Subject Sample	# 3	0.248	12:56
Breath Volume		1.984	12:56
Air Blank		0.000	12:57
Control Test		0.075	12:58
Air Blank		0.000	12:58
Diagnostics Check		OK	12:58

Data Download Date/Time: 10/20/2020 08:54

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAAVILAINEN,JOSEPH,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: OSTERTAG,SHELBY
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 09/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 19:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:48

SUBJECT'S LAST NAME: RODRIGUEZ VAZQUEZ

DATE/TIME OF TEST: 10/09/2020 20:16:03

Diagnostics Check		OK	20:08
Air Blank		0.000	20:09
Control Test		0.079	20:09
Air Blank		0.000	20:09
Subject Sample	# 1	0.220	20:11
Breath Volume		3.023	20:11
Air Blank		0.000	20:11
Air Blank		0.000	20:13
Subject Sample	# 2	0.216	20:14
Breath Volume		2.960	20:14
Air Blank		0.000	20:15
Control Test		0.078	20:15
Air Blank		0.000	20:15
Diagnostics Check		OK	20:16

Data Download Date/Time: 10/20/2020 08:54

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRERO,FRAICOR,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TERRERO,FRAICOR
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 09/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: HINTON

DATE/TIME OF TEST: 10/13/2020 03:34:50

Diagnostics Check		OK	03:28
Air Blank		0.000	03:28
Control Test		0.079	03:29
Air Blank		0.000	03:29
Subject Sample	# 1	0.181	03:30
Breath Volume		1.777	03:30
Air Blank		0.000	03:30
Air Blank		0.000	03:32
Subject Sample	# 2	0.189	03:33
Breath Volume		1.183	03:33
Air Blank		0.000	03:33
Control Test		0.078	03:34
Air Blank		0.000	03:34
Diagnostics Check		OK	03:34

Data Download Date/Time: 10/20/2020 08:54

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,CALVIN,D
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: WALKER,CALVIN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 11:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:00

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 10/13/2020 13:48:32

Diagnostics Check	OK	13:47	
Air Blank	0.000	13:47	
Control Test	0.079	13:48	RFI Detect
Air Blank	0.000	13:48	

Data Download Date/Time: 10/20/2020 08:54

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,CALVIN,D
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: WALKER,CALVIN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 11:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:00

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 10/13/2020 13:56:15

Diagnostics Check		OK	13:52	
Air Blank		0.000	13:52	
Control Test		0.078	13:52	
Air Blank		0.000	13:53	
Subject Sample	# 1		13:54	Subject Test Refused
Breath Volume		0.000	13:54	
Air Blank		0.000	13:55	
Control Test		0.079	13:55	
Air Blank		0.000	13:56	
Diagnostics Check		OK	13:56	

Data Download Date/Time: 10/20/2020 08:54

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRERO,FRAICOR,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TERRERO,FRAICOR
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 09/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: HENRIQUEZ SOLORZANO

DATE/TIME OF TEST: 10/18/2020 00:17:24

Diagnostics Check		OK	00:11
Air Blank		0.000	00:11
Control Test		0.079	00:11
Air Blank		0.000	00:12
Subject Sample	# 1	0.148	00:12
Breath Volume		2.488	00:12
Air Blank		0.000	00:13
Air Blank		0.000	00:15
Subject Sample	# 2	0.140	00:15
Breath Volume		2.019	00:15
Air Blank		0.000	00:16
Control Test		0.078	00:16
Air Blank		0.000	00:17
Diagnostics Check		OK	00:17

Data Download Date/Time: 10/20/2020 08:54

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADENTON,CODY,R
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BARRETT,CODY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 09/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: NEGRON

DATE/TIME OF TEST: 10/18/2020 02:39:57

Diagnostics Check		OK	02:27	
Air Blank		0.000	02:28	
Control Test		0.078	02:28	
Air Blank		0.000	02:28	
Subject Sample	# 1		02:32	No Sample Provided
Breath Volume		0.000	02:32	
Air Blank		0.000	02:32	
Air Blank		0.000	02:34	
Subject Sample	# 2	0.165	02:35	
Breath Volume		2.867	02:35	
Air Blank		0.000	02:35	
Air Blank		0.000	02:37	
Subject Sample	# 3	0.169	02:38	
Breath Volume		3.273	02:38	
Air Blank		0.000	02:39	
Control Test		0.076	02:39	
Air Blank		0.000	02:39	
Diagnostics Check		OK	02:39	

Data Download Date/Time: 10/20/2020 08:54

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRERO,FRAICOR,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TERRERO,FRAICOR
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 09/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: MULLIGAN

DATE/TIME OF TEST: 10/18/2020 04:37:36

Diagnostics Check		OK	04:29
Air Blank		0.000	04:30
Control Test		0.079	04:30
Air Blank		0.000	04:30
Subject Sample	# 1	0.152	04:31
Breath Volume		1.425	04:31
Air Blank		0.000	04:32
Air Blank		0.000	04:33
Subject Sample	# 2	0.161	04:35
Breath Volume		2.445	04:35
Air Blank		0.000	04:36
Control Test		0.078	04:36
Air Blank		0.000	04:37
Diagnostics Check		OK	04:37

Data Download Date/Time: 10/20/2020 08:54

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 10/18/2020 22:16:09

Diagnostics Check		OK	22:08
Air Blank		0.000	22:08
Control Test		0.078	22:09
Air Blank		0.000	22:09
Subject Sample	# 1	0.294	22:11
Breath Volume		3.402	22:11
Air Blank		0.000	22:11
Air Blank		0.000	22:13
Subject Sample	# 2	0.289	22:14
Breath Volume		1.574	22:14
Air Blank		0.000	22:15
Control Test		0.076	22:15
Air Blank		0.000	22:16
Diagnostics Check		OK	22:16

Data Download Date/Time: 10/20/2020 08:54

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRERO,FRAICOR,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TERRERO,FRAICOR
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 09/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: GEISLER

DATE/TIME OF TEST: 10/18/2020 23:18:09

Diagnostics Check		OK	23:02	
Air Blank		0.000	23:03	
Control Test		0.078	23:03	
Air Blank		0.000	23:04	
Subject Sample	# 1	0.000	23:07	Volume Not Met
Breath Volume		0.914	23:07	
Air Blank		0.000	23:07	
Air Blank		0.000	23:09	
Subject Sample	# 2	0.000	23:11	
Breath Volume		1.734	23:11	
Air Blank		0.000	23:11	
Air Blank		0.000	23:13	
Subject Sample	# 3	0.000	23:16	Volume Not Met
Breath Volume		0.484	23:16	
Air Blank		0.000	23:17	
Control Test		0.077	23:17	
Air Blank		0.000	23:18	
Diagnostics Check		OK	23:18	

Data Download Date/Time: 10/20/2020 08:54

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

INSTRUMENT REGISTERED TO: HOLMES BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HURT,ALEXANDER,C
BREATH TEST OPERATOR AGENCY: Holmes Beach PD
ARRESTING OFFICER: HURT,ALEXANDER
ARRESTING OFFICER AGENCY: Holmes Beach PD
LAST AGENCY INSPECTION: 09/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175970
DRY GAS EXPIRY DATE: 08/02/2022
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: CONACHEN

DATE/TIME OF TEST: 10/04/2020 03:29:37

Diagnostics Check		OK	03:22
Air Blank		0.000	03:22
Control Test		0.077	03:23
Air Blank		0.000	03:23
Subject Sample	# 1	0.186	03:24
Breath Volume		1.148	03:24
Air Blank		0.000	03:25
Air Blank		0.000	03:26
Subject Sample	# 2	0.205	03:27
Breath Volume		2.765	03:27
Air Blank		0.000	03:28
Control Test		0.077	03:29
Air Blank		0.000	03:29
Diagnostics Check		OK	03:29

Data Download Date/Time: 10/20/2020 08:55

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

INSTRUMENT REGISTERED TO: HOLMES BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANHORN,MICHAEL,T
BREATH TEST OPERATOR AGENCY: Holmes Beach PD
ARRESTING OFFICER: VANHORN,MICHAEL
ARRESTING OFFICER AGENCY: Holmes Beach PD
LAST AGENCY INSPECTION: 09/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175970
DRY GAS EXPIRY DATE: 08/02/2022
VIOLATION CODE: DUI
ARREST TIME: 05:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:50

SUBJECT'S LAST NAME: SERRANO ORTIZ

DATE/TIME OF TEST: 10/12/2020 06:33:11

Diagnostics Check		OK	06:24
Air Blank		0.000	06:24
Control Test		0.078	06:25
Air Blank		0.000	06:25
Subject Sample	# 1	0.168	06:28
Breath Volume		2.425	06:28
Air Blank		0.000	06:29
Air Blank		0.000	06:30
Subject Sample	# 2	0.166	06:31
Breath Volume		1.792	06:31
Air Blank		0.000	06:32
Control Test		0.077	06:32
Air Blank		0.000	06:33
Diagnostics Check		OK	06:33

Data Download Date/Time: 10/20/2020 08:56

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

INSTRUMENT REGISTERED TO: HOLMES BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HURT,ALEXANDER,C
BREATH TEST OPERATOR AGENCY: Holmes Beach PD
ARRESTING OFFICER: HURT,ALEXANDER
ARRESTING OFFICER AGENCY: Holmes Beach PD
LAST AGENCY INSPECTION: 09/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175970
DRY GAS EXPIRY DATE: 08/02/2022
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: WARD

DATE/TIME OF TEST: 10/17/2020 02:53:52

Diagnostics Check		OK	02:51	
Air Blank		0.000	02:51	
Control Test		0.078	02:51	
Air Blank		0.000	02:52	
Subject Sample	# 1		02:52	Subject Test Refused
Breath Volume		0.000	02:52	
Air Blank		0.000	02:52	
Control Test		0.078	02:53	
Air Blank		0.000	02:53	
Diagnostics Check		OK	02:53	

Data Download Date/Time: 10/20/2020 08:56

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

INSTRUMENT REGISTERED TO: HOLMES BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HURT,ALEXANDER,C
BREATH TEST OPERATOR AGENCY: Holmes Beach PD
ARRESTING OFFICER: HURT,ALEXANDER
ARRESTING OFFICER AGENCY: Holmes Beach PD
LAST AGENCY INSPECTION: 10/20/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175970
DRY GAS EXPIRY DATE: 08/02/2022
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: FALTON

DATE/TIME OF TEST: 10/31/2020 00:31:32

Diagnostics Check		OK	00:24
Air Blank		0.000	00:25
Control Test		0.078	00:25
Air Blank		0.000	00:26
Subject Sample	# 1	0.286	00:26
Breath Volume		2.023	00:26
Air Blank		0.000	00:27
Air Blank		0.000	00:29
Subject Sample	# 2	0.279	00:29
Breath Volume		2.273	00:29
Air Blank		0.000	00:30
Control Test		0.076	00:30
Air Blank		0.000	00:31
Diagnostics Check		OK	00:31

Data Download Date/Time: 11/23/2020 11:13

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 09/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 19:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:51

SUBJECT'S LAST NAME: KELLER

DATE/TIME OF TEST: 10/10/2020 20:16:01

Diagnostics Check	OK	20:14	
Air Blank	0.000	20:15	
Control Test	0.000	20:15	Control Outside Tolerance
Air Blank	0.000	20:16	

Data Download Date/Time: 10/19/2020 09:33

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 09/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 19:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:51

SUBJECT'S LAST NAME: KELLER

DATE/TIME OF TEST: 10/10/2020 20:18:30

Diagnostics Check	OK	20:17	
Air Blank	0.000	20:17	
Control Test	0.000	20:17	Control Outside Tolerance
Air Blank	0.000	20:18	

Data Download Date/Time: 10/19/2020 09:33

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 21:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:47

SUBJECT'S LAST NAME: SUAZO MONTES

DATE/TIME OF TEST: 10/23/2020 22:17:19

Diagnostics Check		OK	22:10
Air Blank		0.000	22:10
Control Test		0.079	22:10
Air Blank		0.000	22:11
Subject Sample	# 1	0.176	22:12
Breath Volume		1.343	22:12
Air Blank		0.000	22:13
Air Blank		0.000	22:15
Subject Sample	# 2	0.193	22:15
Breath Volume		1.523	22:15
Air Blank		0.000	22:16
Control Test		0.077	22:16
Air Blank		0.000	22:17
Diagnostics Check		OK	22:17

Data Download Date/Time: 11/23/2020 12:02

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 04:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:46

SUBJECT'S LAST NAME: MICK

DATE/TIME OF TEST: 10/24/2020 05:13:25

Diagnostics Check		OK	05:06
Air Blank		0.000	05:07
Control Test		0.079	05:07
Air Blank		0.000	05:08
Subject Sample	# 1	0.162	05:08
Breath Volume		1.382	05:08
Air Blank		0.000	05:09
Air Blank		0.000	05:11
Subject Sample	# 2	0.168	05:11
Breath Volume		2.195	05:11
Air Blank		0.000	05:12
Control Test		0.078	05:12
Air Blank		0.000	05:13
Diagnostics Check		OK	05:13

Data Download Date/Time: 11/23/2020 12:02

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: SYS CHECK
ARREST TIME: 23:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 10/28/2020 23:43:42

Diagnostics Check		OK	23:41	
Air Blank		0.000	23:41	
Control Test		0.076	23:41	
Air Blank		0.000	23:42	
Subject Sample	# 1		23:42	Interferent Detect
Breath Volume		2.085	23:42	
Air Blank			23:43	Ambient Fail
Air Blank		0.000	23:43	

Data Download Date/Time: 11/23/2020 12:02

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: SYS CHECK
ARREST TIME: 23:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 10/28/2020 23:59:16

Diagnostics Check		OK	23:53
Air Blank		0.000	23:53
Control Test		0.076	23:53
Air Blank		0.000	23:54
Subject Sample	# 1	0.000	23:55
Breath Volume		2.335	23:55
Air Blank		0.000	23:55
Air Blank		0.000	23:57
Subject Sample	# 2	0.000	23:57
Breath Volume		2.058	23:57
Air Blank		0.000	23:58
Control Test		0.078	23:58
Air Blank		0.000	23:59
Diagnostics Check		OK	23:59

Data Download Date/Time: 11/23/2020 12:02

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006760

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUBEL,NEIL,F
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RUBEL,NEIL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 16:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:08

SUBJECT'S LAST NAME: MOSHER

DATE/TIME OF TEST: 10/30/2020 17:35:47

Diagnostics Check		OK	17:28
Air Blank		0.000	17:29
Control Test		0.080	17:29
Air Blank		0.000	17:30
Subject Sample	# 1	0.180	17:30
Breath Volume		2.042	17:30
Air Blank		0.000	17:31
Air Blank		0.000	17:33
Subject Sample	# 2	0.178	17:34
Breath Volume		2.390	17:34
Air Blank		0.000	17:34
Control Test		0.080	17:35
Air Blank		0.000	17:35
Diagnostics Check		OK	17:35

Data Download Date/Time: 11/05/2020 11:41

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 10/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 10/03/2020 22:27:07

Diagnostics Check		OK	22:20
Air Blank		0.000	22:20
Control Test		0.078	22:21
Air Blank		0.000	22:21
Subject Sample	# 1	0.139	22:22
Breath Volume		2.980	22:22
Air Blank		0.000	22:23
Air Blank		0.000	22:24
Subject Sample	# 2	0.144	22:25
Breath Volume		4.179	22:25
Air Blank		0.000	22:26
Control Test		0.078	22:26
Air Blank		0.000	22:27
Diagnostics Check		OK	22:27

Data Download Date/Time: 11/09/2020 10:53

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 10/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 20:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: DAVIDSON

DATE/TIME OF TEST: 10/04/2020 21:35:23

Diagnostics Check		OK	21:29
Air Blank		0.000	21:29
Control Test		0.079	21:30
Air Blank		0.000	21:30
Subject Sample	# 1	0.262	21:30
Breath Volume		4.289	21:30
Air Blank		0.000	21:31
Air Blank		0.000	21:33
Subject Sample	# 2	0.252	21:33
Breath Volume		3.042	21:33
Air Blank		0.000	21:34
Control Test		0.075	21:34
Air Blank		0.000	21:35
Diagnostics Check		OK	21:35

Data Download Date/Time: 11/09/2020 10:53

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 10/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: YODER

DATE/TIME OF TEST: 10/17/2020 02:19:25

Diagnostics Check		OK	02:13
Air Blank		0.000	02:13
Control Test		0.078	02:13
Air Blank		0.000	02:14
Subject Sample	# 1	0.174	02:14
Breath Volume		4.535	02:14
Air Blank		0.000	02:15
Air Blank		0.000	02:17
Subject Sample	# 2	0.171	02:17
Breath Volume		4.636	02:17
Air Blank		0.000	02:18
Control Test		0.079	02:18
Air Blank		0.000	02:19
Diagnostics Check		OK	02:19

Data Download Date/Time: 11/09/2020 10:53

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNEH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 10/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 10/18/2020 02:44:47

Diagnostics Check		OK	02:41	
Air Blank		0.000	02:42	
Control Test		0.080	02:42	
Air Blank		0.000	02:42	
Subject Sample	# 1		02:43	Subject Test Refused
Breath Volume		0.000	02:43	
Air Blank		0.000	02:43	
Control Test		0.078	02:44	
Air Blank		0.000	02:44	
Diagnostics Check		OK	02:44	

Data Download Date/Time: 11/09/2020 10:53

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 10/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 10/22/2020 00:16:54

Diagnostics Check		OK	00:03	
Air Blank		0.000	00:03	
Control Test		0.075	00:03	
Air Blank		0.000	00:04	
Subject Sample	# 1	0.163	00:04	
Breath Volume		2.019	00:04	
Air Blank		0.000	00:05	
Air Blank		0.000	00:07	
Subject Sample	# 2	0.127	00:10	Volume Not Met
Breath Volume		0.109	00:10	
Air Blank		0.000	00:11	
Air Blank		0.000	00:12	
Subject Sample	# 3		00:15	Subject Test Refused
Breath Volume		0.183	00:15	
Air Blank		0.000	00:15	
Control Test		0.073	00:16	Control Outside Tolerance
Air Blank		0.000	00:16	

Data Download Date/Time: 11/09/2020 10:53

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 10/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: MAINWOOD

DATE/TIME OF TEST: 10/23/2020 02:41:17

Diagnostics Check		OK	02:35
Air Blank		0.000	02:35
Control Test		0.078	02:36
Air Blank		0.000	02:36
Subject Sample	# 1	0.126	02:36
Breath Volume		2.800	02:36
Air Blank		0.000	02:37
Air Blank		0.000	02:39
Subject Sample	# 2	0.123	02:39
Breath Volume		2.753	02:39
Air Blank		0.000	02:40
Control Test		0.078	02:40
Air Blank		0.000	02:41
Diagnostics Check		OK	02:41

Data Download Date/Time: 11/09/2020 10:54

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 10/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 10/03/2020 02:28:57

Diagnostics Check		OK	02:22
Air Blank		0.000	02:23
Control Test		0.080	02:23
Air Blank		0.000	02:24
Subject Sample	# 1	0.150	02:24
Breath Volume		4.187	02:24
Air Blank		0.000	02:25
Air Blank		0.000	02:26
Subject Sample	# 2	0.150	02:27
Breath Volume		3.628	02:27
Air Blank		0.000	02:28
Control Test		0.079	02:28
Air Blank		0.000	02:28
Diagnostics Check		OK	02:28

Data Download Date/Time: 11/04/2020 12:20

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 10/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 10/04/2020 03:24:40

Diagnostics Check		OK	03:17
Air Blank		0.000	03:18
Control Test		0.080	03:18
Air Blank		0.000	03:19
Subject Sample	# 1	0.200	03:19
Breath Volume		1.511	03:19
Air Blank		0.000	03:20
Air Blank		0.000	03:22
Subject Sample	# 2	0.197	03:23
Breath Volume		2.230	03:23
Air Blank		0.000	03:23
Control Test		0.080	03:24
Air Blank		0.000	03:24
Diagnostics Check		OK	03:24

Data Download Date/Time: 11/04/2020 12:20

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 10/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: MCGIBANY

DATE/TIME OF TEST: 10/08/2020 02:38:34

Diagnostics Check		OK	02:32
Air Blank		0.000	02:32
Control Test		0.080	02:33
Air Blank		0.000	02:33
Subject Sample	# 1	0.101	02:34
Breath Volume		3.281	02:34
Air Blank		0.000	02:34
Air Blank		0.000	02:36
Subject Sample	# 2	0.103	02:37
Breath Volume		2.925	02:37
Air Blank		0.000	02:37
Control Test		0.078	02:37
Air Blank		0.000	02:38
Diagnostics Check		OK	02:38

Data Download Date/Time: 11/04/2020 12:20

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 10/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 10/10/2020 02:05:25

Diagnostics Check		OK	01:58
Air Blank		0.000	01:59
Control Test		0.079	01:59
Air Blank		0.000	02:00
Subject Sample	# 1	0.216	02:00
Breath Volume		2.175	02:00
Air Blank		0.000	02:01
Air Blank		0.000	02:03
Subject Sample	# 2	0.217	02:03
Breath Volume		2.500	02:03
Air Blank		0.000	02:04
Control Test		0.078	02:04
Air Blank		0.000	02:05
Diagnostics Check		OK	02:05

Data Download Date/Time: 11/04/2020 12:20

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 10/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: CASON

DATE/TIME OF TEST: 10/11/2020 02:33:34

Diagnostics Check		OK	02:25
Air Blank		0.000	02:26
Control Test		0.078	02:26
Air Blank		0.000	02:27
Subject Sample	# 1	0.189	02:28
Breath Volume		1.621	02:28
Air Blank		0.000	02:28
Air Blank		0.000	02:30
Subject Sample	# 2	0.184	02:31
Breath Volume		1.398	02:31
Air Blank		0.000	02:32
Control Test		0.078	02:32
Air Blank		0.000	02:33
Diagnostics Check		OK	02:33

Data Download Date/Time: 11/04/2020 12:20

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 10/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: WHEELER

DATE/TIME OF TEST: 10/23/2020 01:03:35

Diagnostics Check		OK	00:57
Air Blank		0.000	00:57
Control Test		0.079	00:58
Air Blank		0.000	00:58
Subject Sample	# 1	0.084	00:59
Breath Volume		2.441	00:59
Air Blank		0.000	00:59
Air Blank		0.000	01:01
Subject Sample	# 2	0.085	01:02
Breath Volume		2.757	01:02
Air Blank		0.000	01:02
Control Test		0.078	01:02
Air Blank		0.000	01:03
Diagnostics Check		OK	01:03

Data Download Date/Time: 11/04/2020 12:20

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 10/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: TRUESDALE

DATE/TIME OF TEST: 10/24/2020 02:13:08

Diagnostics Check		OK	02:06
Air Blank		0.000	02:07
Control Test		0.079	02:07
Air Blank		0.000	02:08
Subject Sample	# 1	0.147	02:08
Breath Volume		1.406	02:08
Air Blank		0.000	02:09
Air Blank		0.000	02:11
Subject Sample	# 2	0.144	02:11
Breath Volume		2.390	02:11
Air Blank		0.000	02:12
Control Test		0.079	02:12
Air Blank		0.000	02:13
Diagnostics Check		OK	02:13

Data Download Date/Time: 11/04/2020 12:20

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANCO,G,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: FRANCO,G
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: VEGA

DATE/TIME OF TEST: 10/02/2020 03:20:27

Diagnostics Check		OK	03:17	
Air Blank		0.000	03:17	
Control Test		0.076	03:18	
Air Blank		0.000	03:18	
Subject Sample	# 1		03:19	Subject Test Refused
Breath Volume		0.000	03:19	
Air Blank		0.000	03:19	
Control Test		0.076	03:19	
Air Blank		0.000	03:20	
Diagnostics Check		OK	03:20	

Data Download Date/Time: 10/28/2020 01:57

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: GONZALEZ MOREJON

DATE/TIME OF TEST: 10/03/2020 01:55:30

Diagnostics Check		OK	01:52	
Air Blank		0.000	01:53	
Control Test		0.076	01:53	
Air Blank		0.000	01:53	
Subject Sample	# 1		01:54	Subject Test Refused
Breath Volume		0.000	01:54	
Air Blank		0.000	01:54	
Control Test		0.076	01:54	
Air Blank		0.000	01:55	
Diagnostics Check		OK	01:55	

Data Download Date/Time: 10/28/2020 01:57

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: MELENDEZ,D
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: ALVAREZ REYES

DATE/TIME OF TEST: 10/03/2020 04:45:12

Diagnostics Check		OK	04:42	
Air Blank		0.000	04:42	
Control Test		0.078	04:43	
Air Blank		0.000	04:43	
Subject Sample	# 1		04:43	Subject Test Refused
Breath Volume		0.000	04:43	
Air Blank		0.000	04:44	
Control Test		0.077	04:44	
Air Blank		0.000	04:45	
Diagnostics Check		OK	04:45	

Data Download Date/Time: 10/28/2020 01:57

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,F
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: DALY

DATE/TIME OF TEST: 10/04/2020 02:51:08

Diagnostics Check		OK	02:48	
Air Blank		0.000	02:48	
Control Test		0.077	02:49	
Air Blank		0.000	02:49	
Subject Sample	# 1		02:49	Subject Test Refused
Breath Volume		0.000	02:49	
Air Blank		0.000	02:50	
Control Test		0.076	02:50	
Air Blank		0.000	02:51	
Diagnostics Check		OK	02:51	

Data Download Date/Time: 10/28/2020 01:57

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 06:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:10

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 10/07/2020 06:39:47

Diagnostics Check		OK	06:33
Air Blank		0.000	06:33
Control Test		0.078	06:34
Air Blank		0.000	06:34
Subject Sample	# 1	0.000	06:35
Breath Volume		1.980	06:35
Air Blank		0.000	06:35
Air Blank		0.000	06:37
Subject Sample	# 2	0.000	06:38
Breath Volume		2.113	06:38
Air Blank		0.000	06:38
Control Test		0.078	06:39
Air Blank		0.000	06:39
Diagnostics Check		OK	06:39

Data Download Date/Time: 10/28/2020 01:57

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:02

SUBJECT'S LAST NAME: DUNGAN

DATE/TIME OF TEST: 10/08/2020 06:38:12

Diagnostics Check		OK	06:31
Air Blank		0.000	06:32
Control Test		0.077	06:32
Air Blank		0.000	06:32
Subject Sample	# 1	0.181	06:33
Breath Volume		1.585	06:33
Air Blank		0.000	06:34
Air Blank		0.000	06:36
Subject Sample	# 2	0.184	06:36
Breath Volume		1.203	06:36
Air Blank		0.000	06:37
Control Test		0.077	06:37
Air Blank		0.000	06:38
Diagnostics Check		OK	06:38

Data Download Date/Time: 10/28/2020 01:57

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: HIBBERT

DATE/TIME OF TEST: 10/11/2020 00:41:45

Diagnostics Check		OK	00:30	
Air Blank		0.000	00:30	
Control Test		0.077	00:31	
Air Blank		0.000	00:31	
Subject Sample	# 1	0.165	00:34	Volume Not Met
Breath Volume		0.300	00:34	
Air Blank		0.000	00:35	
Air Blank		0.000	00:37	
Subject Sample	# 2		00:40	Subject Test Refused
Breath Volume		0.367	00:40	
Air Blank		0.000	00:40	
Control Test		0.075	00:41	
Air Blank		0.000	00:41	
Diagnostics Check		OK	00:41	

Data Download Date/Time: 10/28/2020 01:57

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,F
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 10/17/2020 04:02:43

Diagnostics Check		OK	03:59	
Air Blank		0.000	04:00	
Control Test		0.077	04:00	
Air Blank		0.000	04:01	
Subject Sample	# 1		04:01	Subject Test Refused
Breath Volume		0.000	04:01	
Air Blank		0.000	04:01	
Control Test		0.076	04:02	
Air Blank		0.000	04:02	
Diagnostics Check		OK	04:02	

Data Download Date/Time: 10/28/2020 01:57

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PACHECO,GEORGE,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: PACHECO,GEORGE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 10/23/2020 01:32:30

Diagnostics Check		OK	01:28	
Air Blank		0.000	01:29	
Control Test		0.077	01:29	
Air Blank		0.000	01:29	
Subject Sample	# 1		01:31	Subject Test Refused
Breath Volume		0.000	01:31	
Air Blank		0.000	01:31	
Control Test		0.077	01:31	
Air Blank		0.000	01:32	
Diagnostics Check		OK	01:32	

Data Download Date/Time: 10/28/2020 01:57

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANCO,G,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: FRANCO,G
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: HERNANDEZ SARMIENTO

DATE/TIME OF TEST: 10/24/2020 02:44:43

Diagnostics Check		OK	02:41	
Air Blank		0.000	02:42	
Control Test		0.078	02:42	
Air Blank		0.000	02:43	
Subject Sample	# 1		02:43	Subject Test Refused
Breath Volume		0.000	02:43	
Air Blank		0.000	02:43	
Control Test		0.077	02:44	
Air Blank		0.000	02:44	
Diagnostics Check		OK	02:44	

Data Download Date/Time: 10/28/2020 01:57

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:03

SUBJECT'S LAST NAME: MCCUTCHEN

DATE/TIME OF TEST: 10/24/2020 06:39:36

Diagnostics Check		OK	06:36	
Air Blank		0.000	06:37	
Control Test		0.078	06:37	
Air Blank		0.000	06:38	
Subject Sample	# 1		06:38	Subject Test Refused
Breath Volume		0.000	06:38	
Air Blank		0.000	06:38	
Control Test		0.077	06:39	
Air Blank		0.000	06:39	
Diagnostics Check		OK	06:39	

Data Download Date/Time: 10/28/2020 01:57

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: QUINTANA,J
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: MARQUINA

DATE/TIME OF TEST: 10/25/2020 03:11:52

Diagnostics Check		OK	03:08	
Air Blank		0.000	03:09	
Control Test		0.077	03:09	
Air Blank		0.000	03:10	
Subject Sample	# 1		03:10	Subject Test Refused
Breath Volume		0.000	03:10	
Air Blank		0.000	03:10	
Control Test		0.077	03:11	
Air Blank		0.000	03:11	
Diagnostics Check		OK	03:11	

Data Download Date/Time: 10/28/2020 01:57

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: VELEZ,JONATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 10/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 20:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:39

SUBJECT'S LAST NAME: MESTRE

DATE/TIME OF TEST: 10/08/2020 22:10:49

Diagnostics Check		OK	21:59	
Air Blank		0.000	21:59	
Control Test		0.079	22:00	
Air Blank		0.000	22:00	
Subject Sample	# 1	0.000	22:03	Volume Not Met
Breath Volume		0.292	22:03	
Air Blank		0.000	22:04	
Air Blank		0.000	22:06	
Subject Sample	# 2		22:09	No Sample Provided
Breath Volume		0.000	22:09	
Air Blank		0.000	22:09	
Control Test		0.078	22:10	
Air Blank		0.000	22:10	
Diagnostics Check		OK	22:10	

Data Download Date/Time: 11/23/2020 14:38

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: VELEZ,JONATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 10/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 20:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:39

SUBJECT'S LAST NAME: MESTRE

DATE/TIME OF TEST: 10/08/2020 22:28:34

Diagnostics Check		OK	22:13	
Air Blank		0.000	22:13	
Control Test		0.079	22:13	
Air Blank		0.000	22:14	
Subject Sample	# 1	0.000	22:17	Volume Not Met
Breath Volume		0.082	22:17	
Air Blank		0.000	22:18	
Air Blank		0.000	22:20	
Subject Sample	# 2	0.000	22:21	
Breath Volume		1.226	22:21	
Air Blank		0.000	22:22	
Air Blank		0.000	22:24	
Subject Sample	# 3	0.000	22:27	Volume Not Met
Breath Volume		0.800	22:27	
Air Blank		0.000	22:27	
Control Test		0.079	22:27	
Air Blank		0.000	22:28	
Diagnostics Check		OK	22:28	

Data Download Date/Time: 11/23/2020 14:38

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ROJAS,RANDY
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 10/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: STREMINSKY

DATE/TIME OF TEST: 10/10/2020 02:44:32

Diagnostics Check		OK	02:38
Air Blank		0.000	02:38
Control Test		0.079	02:38
Air Blank		0.000	02:39
Subject Sample	# 1	0.000	02:39
Breath Volume		3.566	02:39
Air Blank		0.000	02:40
Air Blank		0.000	02:42
Subject Sample	# 2	0.000	02:43
Breath Volume		3.773	02:43
Air Blank		0.000	02:43
Control Test		0.080	02:43
Air Blank		0.000	02:44
Diagnostics Check		OK	02:44

Data Download Date/Time: 11/23/2020 14:39

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAPP,JOSHUA,R
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ROJAS,R
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 10/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 06:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:07

SUBJECT'S LAST NAME: LEVY

DATE/TIME OF TEST: 10/10/2020 07:42:36

Diagnostics Check		OK	07:36
Air Blank		0.000	07:36
Control Test		0.080	07:37
Air Blank		0.000	07:37
Subject Sample	# 1	0.000	07:38
Breath Volume		2.707	07:38
Air Blank		0.000	07:38
Air Blank		0.000	07:40
Subject Sample	# 2	0.000	07:41
Breath Volume		2.472	07:41
Air Blank		0.000	07:41
Control Test		0.079	07:41
Air Blank		0.000	07:42
Diagnostics Check		OK	07:42

Data Download Date/Time: 11/23/2020 14:39

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CAMPBELL,DWAYNE
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 10/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 19:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:07

SUBJECT'S LAST NAME: BROWN JR

DATE/TIME OF TEST: 10/10/2020 20:40:50

Diagnostics Check		OK	20:34
Air Blank		0.000	20:35
Control Test		0.080	20:35
Air Blank		0.000	20:35
Subject Sample	# 1	0.144	20:36
Breath Volume		3.179	20:36
Air Blank		0.000	20:37
Air Blank		0.000	20:38
Subject Sample	# 2	0.145	20:39
Breath Volume		2.460	20:39
Air Blank		0.000	20:39
Control Test		0.079	20:40
Air Blank		0.000	20:40
Diagnostics Check		OK	20:40

Data Download Date/Time: 11/23/2020 14:39

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MCGUIRE,THOMAS
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 10/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 10/10/2020 23:01:43

Diagnostics Check		OK	22:54
Air Blank		0.000	22:54
Control Test		0.079	22:55
Air Blank		0.000	22:55
Subject Sample	# 1	0.203	22:56
Breath Volume		1.781	22:56
Air Blank		0.000	22:57
Air Blank		0.000	22:58
Subject Sample	# 2	0.203	22:59
Breath Volume		1.601	22:59
Air Blank		0.000	23:00
Control Test		0.079	23:01
Air Blank		0.000	23:01
Diagnostics Check		OK	23:01

Data Download Date/Time: 11/23/2020 14:39

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BEAURGARD,JEFFREY
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 10/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: KATZ

DATE/TIME OF TEST: 10/11/2020 00:39:15

Diagnostics Check		OK	00:33
Air Blank		0.000	00:33
Control Test		0.079	00:33
Air Blank		0.000	00:34
Subject Sample	# 1	0.000	00:34
Breath Volume		3.804	00:34
Air Blank		0.000	00:35
Air Blank		0.000	00:37
Subject Sample	# 2	0.000	00:37
Breath Volume		3.183	00:37
Air Blank		0.000	00:38
Control Test		0.079	00:38
Air Blank		0.000	00:39
Diagnostics Check		OK	00:39

Data Download Date/Time: 11/23/2020 14:39

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HOLLIGAN,DANIEL
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 10/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: ARIAS CAMPO

DATE/TIME OF TEST: 10/13/2020 03:43:30

Diagnostics Check		OK	03:36
Air Blank		0.000	03:36
Control Test		0.079	03:36
Air Blank		0.000	03:37
Subject Sample	# 1	0.233	03:38
Breath Volume		1.574	03:38
Air Blank		0.000	03:39
Air Blank		0.000	03:40
Subject Sample	# 2	0.238	03:41
Breath Volume		1.511	03:41
Air Blank		0.000	03:42
Control Test		0.079	03:42
Air Blank		0.000	03:43
Diagnostics Check		OK	03:43

Data Download Date/Time: 11/23/2020 14:39

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: PRESUME,JEHUD
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 10/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:37

SUBJECT'S LAST NAME: DORELUS

DATE/TIME OF TEST: 10/16/2020 22:11:31

Diagnostics Check		OK	22:02
Air Blank		0.000	22:03
Control Test		0.079	22:03
Air Blank		0.000	22:04
Subject Sample	# 1	0.187	22:06
Breath Volume		1.406	22:06
Air Blank		0.000	22:07
Air Blank		0.000	22:08
Subject Sample	# 2	0.184	22:09
Breath Volume		1.281	22:09
Air Blank		0.000	22:10
Control Test		0.075	22:10
Air Blank		0.000	22:11
Diagnostics Check		OK	22:11

Data Download Date/Time: 11/23/2020 14:39

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:43

SUBJECT'S LAST NAME: PORTES

DATE/TIME OF TEST: 10/16/2020 23:17:14

Diagnostics Check		OK	23:10
Air Blank		0.000	23:10
Control Test		0.079	23:11
Air Blank		0.000	23:11
Subject Sample	# 1	0.086	23:12
Breath Volume		1.679	23:12
Air Blank		0.000	23:13
Air Blank		0.000	23:14
Subject Sample	# 2	0.085	23:15
Breath Volume		1.789	23:15
Air Blank		0.000	23:16
Control Test		0.078	23:16
Air Blank		0.000	23:17
Diagnostics Check		OK	23:17

Data Download Date/Time: 11/23/2020 14:39

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 04:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:54

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 10/17/2020 05:21:25

Diagnostics Check		OK	05:15
Air Blank		0.000	05:15
Control Test		0.079	05:15
Air Blank		0.000	05:16
Subject Sample	# 1	0.139	05:16
Breath Volume		1.878	05:16
Air Blank		0.000	05:17
Air Blank		0.000	05:19
Subject Sample	# 2	0.137	05:19
Breath Volume		1.855	05:19
Air Blank		0.000	05:20
Control Test		0.078	05:20
Air Blank		0.000	05:21
Diagnostics Check		OK	05:21

Data Download Date/Time: 11/23/2020 14:39

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: AKERS,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 06:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:30

SUBJECT'S LAST NAME: HEREDIA MORAN

DATE/TIME OF TEST: 10/18/2020 08:12:32

Diagnostics Check		OK	08:04
Air Blank		0.000	08:04
Control Test		0.080	08:05
Air Blank		0.000	08:05
Subject Sample	# 1	0.190	08:06
Breath Volume		1.343	08:06
Air Blank		0.000	08:06
Air Blank		0.000	08:08
Subject Sample	# 2	0.177	08:10
Breath Volume		1.230	08:10
Air Blank		0.000	08:11
Control Test		0.079	08:11
Air Blank		0.000	08:12
Diagnostics Check		OK	08:12

Data Download Date/Time: 11/23/2020 14:40

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: PRESUME,JEHUD
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 10/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: GERARD

DATE/TIME OF TEST: 10/20/2020 00:16:22

Diagnostics Check		OK	00:09
Air Blank		0.000	00:09
Control Test		0.079	00:09
Air Blank		0.000	00:10
Subject Sample	# 1	0.210	00:11
Breath Volume		1.644	00:11
Air Blank		0.000	00:12
Air Blank		0.000	00:13
Subject Sample	# 2	0.196	00:14
Breath Volume		1.875	00:14
Air Blank		0.000	00:15
Control Test		0.079	00:15
Air Blank		0.000	00:16
Diagnostics Check		OK	00:16

Data Download Date/Time: 11/23/2020 14:40

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: YOSHIHITO

DATE/TIME OF TEST: 10/21/2020 01:15:05

Diagnostics Check		OK	01:07
Air Blank		0.000	01:07
Control Test		0.080	01:07
Air Blank		0.000	01:08
Subject Sample	# 1	0.207	01:09
Breath Volume		1.796	01:09
Air Blank		0.000	01:10
Air Blank		0.000	01:12
Subject Sample	# 2	0.196	01:13
Breath Volume		1.386	01:13
Air Blank		0.000	01:14
Control Test		0.075	01:14
Air Blank		0.000	01:15
Diagnostics Check		OK	01:15

Data Download Date/Time: 11/23/2020 14:40

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MONTE DE OCA
VIERA,ALEXANDER
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 10/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 19:31
GENDER: Male

OBSERVATION PERIOD BEGAN: 20:43

SUBJECT'S LAST NAME: FIGARO

DATE/TIME OF TEST: 10/23/2020 21:10:57

Diagnostics Check		OK	21:04
Air Blank		0.000	21:05
Control Test		0.080	21:05
Air Blank		0.000	21:06
Subject Sample	# 1	0.000	21:06
Breath Volume		2.730	21:06
Air Blank		0.000	21:07
Air Blank		0.000	21:09
Subject Sample	# 2	0.000	21:09
Breath Volume		2.484	21:09
Air Blank		0.000	21:10
Control Test		0.079	21:10
Air Blank		0.000	21:10
Diagnostics Check		OK	21:10

Data Download Date/Time: 11/23/2020 14:40

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: GONCALVES DEMENDONCA

DATE/TIME OF TEST: 10/24/2020 01:37:37

Diagnostics Check		OK	01:27
Air Blank		0.000	01:27
Control Test		0.079	01:28
Air Blank		0.000	01:28
Subject Sample	# 1	0.116	01:31
Breath Volume		2.035	01:31
Air Blank		0.000	01:31
Air Blank		0.000	01:33
Subject Sample	# 2	0.126	01:35
Breath Volume		2.125	01:35
Air Blank		0.000	01:36
Control Test		0.076	01:36
Air Blank		0.000	01:37
Diagnostics Check		OK	01:37

Data Download Date/Time: 11/23/2020 14:40

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 18:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: CARSON

DATE/TIME OF TEST: 10/27/2020 20:08:20

Diagnostics Check		OK	20:01
Air Blank		0.000	20:02
Control Test		0.079	20:02
Air Blank		0.000	20:03
Subject Sample	# 1	0.178	20:03
Breath Volume		1.324	20:03
Air Blank		0.000	20:04
Air Blank		0.000	20:05
Subject Sample	# 2	0.179	20:06
Breath Volume		2.367	20:06
Air Blank		0.000	20:07
Control Test		0.079	20:07
Air Blank		0.000	20:08
Diagnostics Check		OK	20:08

Data Download Date/Time: 11/23/2020 14:41

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: STACHURA,HAYDEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: LYONS

DATE/TIME OF TEST: 10/30/2020 02:19:30

Diagnostics Check		OK	02:07
Air Blank		0.000	02:07
Control Test		0.080	02:08
Air Blank		0.000	02:08
Subject Sample	# 1	0.238	02:10
Breath Volume		2.175	02:10
Air Blank		0.000	02:11
Air Blank		0.000	02:13
Subject Sample	# 2	0.217	02:14
Breath Volume		2.222	02:14
Air Blank		0.000	02:15
Air Blank		0.000	02:17
Subject Sample	# 3	0.228	02:17
Breath Volume		1.890	02:17
Air Blank		0.000	02:18
Control Test		0.078	02:18
Air Blank		0.000	02:19
Diagnostics Check		OK	02:19

Data Download Date/Time: 11/23/2020 14:41

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: RADANOF,BARRY
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 10/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: KEUNING

DATE/TIME OF TEST: 10/30/2020 03:15:46

Diagnostics Check		OK	03:03	
Air Blank		0.000	03:03	
Control Test		0.078	03:04	
Air Blank		0.000	03:04	
Subject Sample	# 1	0.167	03:07	Volume Not Met
Breath Volume		0.253	03:07	
Air Blank		0.000	03:08	
Air Blank		0.000	03:10	
Subject Sample	# 2	0.180	03:10	
Breath Volume		1.328	03:10	
Air Blank		0.000	03:11	
Air Blank		0.000	03:13	
Subject Sample	# 3	0.172	03:14	
Breath Volume		1.859	03:14	
Air Blank		0.000	03:14	
Control Test		0.078	03:15	
Air Blank		0.000	03:15	
Diagnostics Check		OK	03:15	

Data Download Date/Time: 11/23/2020 14:41

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: MAHARAJ

DATE/TIME OF TEST: 10/31/2020 02:52:25

Diagnostics Check		OK	02:45
Air Blank		0.000	02:46
Control Test		0.080	02:46
Air Blank		0.000	02:47
Subject Sample	# 1	0.263	02:47
Breath Volume		1.863	02:47
Air Blank		0.000	02:48
Air Blank		0.000	02:50
Subject Sample	# 2	0.253	02:50
Breath Volume		2.125	02:50
Air Blank		0.000	02:51
Control Test		0.078	02:51
Air Blank		0.000	02:52
Diagnostics Check		OK	02:52

Data Download Date/Time: 11/23/2020 14:41

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HERNANDEZ,ERNESTO
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 09/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: MOJOCOA

DATE/TIME OF TEST: 10/01/2020 00:50:09

Diagnostics Check		OK	00:43
Air Blank		0.000	00:44
Control Test		0.080	00:44
Air Blank		0.000	00:45
Subject Sample	# 1	0.124	00:45
Breath Volume		2.546	00:45
Air Blank		0.000	00:46
Air Blank		0.000	00:48
Subject Sample	# 2	0.116	00:48
Breath Volume		2.960	00:48
Air Blank		0.000	00:49
Control Test		0.080	00:49
Air Blank		0.000	00:50
Diagnostics Check		OK	00:50

Data Download Date/Time: 10/08/2020 13:04

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 09/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: CRUMLICH

DATE/TIME OF TEST: 10/02/2020 00:45:25

Diagnostics Check		OK	00:38
Air Blank		0.000	00:38
Control Test		0.080	00:39
Air Blank		0.000	00:39
Subject Sample	# 1	0.113	00:40
Breath Volume		1.453	00:40
Air Blank		0.000	00:40
Air Blank		0.000	00:42
Subject Sample	# 2	0.107	00:43
Breath Volume		1.753	00:43
Air Blank		0.000	00:44
Control Test		0.080	00:44
Air Blank		0.000	00:45
Diagnostics Check		OK	00:45

Data Download Date/Time: 10/08/2020 13:04

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SABAN,SHAHEED
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 09/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: LEYVA

DATE/TIME OF TEST: 10/02/2020 03:00:23

Diagnostics Check		OK	02:53
Air Blank		0.000	02:54
Control Test		0.079	02:54
Air Blank		0.000	02:54
Subject Sample	# 1	0.233	02:55
Breath Volume		1.941	02:55
Air Blank		0.000	02:56
Air Blank		0.000	02:58
Subject Sample	# 2	0.240	02:58
Breath Volume		1.175	02:58
Air Blank		0.000	02:59
Control Test		0.078	02:59
Air Blank		0.000	03:00
Diagnostics Check		OK	03:00

Data Download Date/Time: 10/08/2020 13:04

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 09/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 19:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: SQUILLANTE

DATE/TIME OF TEST: 10/02/2020 20:33:08

Diagnostics Check		OK	20:27
Air Blank		0.000	20:27
Control Test		0.080	20:27
Air Blank		0.000	20:28
Subject Sample	# 1	0.000	20:28
Breath Volume		1.531	20:28
Air Blank		0.000	20:29
Air Blank		0.000	20:31
Subject Sample	# 2	0.000	20:31
Breath Volume		1.664	20:31
Air Blank		0.000	20:32
Control Test		0.079	20:32
Air Blank		0.000	20:33
Diagnostics Check		OK	20:33

Data Download Date/Time: 10/08/2020 13:04

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FITHIAN,JASON
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 09/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: OVERTON

DATE/TIME OF TEST: 10/02/2020 22:50:14

Diagnostics Check		OK	22:44
Air Blank		0.000	22:44
Control Test		0.080	22:44
Air Blank		0.000	22:45
Subject Sample	# 1	0.230	22:45
Breath Volume		1.773	22:45
Air Blank		0.000	22:46
Air Blank		0.000	22:48
Subject Sample	# 2	0.229	22:48
Breath Volume		2.203	22:48
Air Blank		0.000	22:49
Control Test		0.078	22:49
Air Blank		0.000	22:50
Diagnostics Check		OK	22:50

Data Download Date/Time: 10/08/2020 13:04

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SELDOMRIDGE,STEVEN
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 09/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: MOATS

DATE/TIME OF TEST: 10/03/2020 00:58:18

Diagnostics Check		OK	00:41	
Air Blank		0.000	00:41	
Control Test		0.080	00:42	
Air Blank		0.000	00:42	
Subject Sample	# 1	0.117	00:45	Volume Not Met
Breath Volume		0.363	00:45	
Air Blank		0.000	00:46	
Air Blank		0.000	00:48	
Subject Sample	# 2	0.106	00:51	
Breath Volume		1.445	00:51	
Air Blank		0.000	00:51	
Air Blank		0.000	00:53	
Subject Sample	# 3	0.112	00:56	
Breath Volume		1.136	00:56	
Air Blank		0.000	00:57	
Control Test		0.078	00:57	
Air Blank		0.000	00:58	
Diagnostics Check		OK	00:58	

Data Download Date/Time: 10/08/2020 13:04

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 09/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: PIZARRO

DATE/TIME OF TEST: 10/03/2020 03:17:33

Diagnostics Check		OK	03:10
Air Blank		0.000	03:10
Control Test		0.080	03:10
Air Blank		0.000	03:11
Subject Sample	# 1	0.146	03:12
Breath Volume		1.558	03:12
Air Blank		0.000	03:13
Air Blank		0.000	03:15
Subject Sample	# 2	0.152	03:15
Breath Volume		2.066	03:15
Air Blank		0.000	03:16
Control Test		0.080	03:16
Air Blank		0.000	03:17
Diagnostics Check		OK	03:17

Data Download Date/Time: 10/08/2020 13:04

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ROJAS,RANDY
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 09/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 04:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:48

SUBJECT'S LAST NAME: ARGUETA

DATE/TIME OF TEST: 10/04/2020 06:24:50

Diagnostics Check		OK	06:18
Air Blank		0.000	06:19
Control Test		0.079	06:19
Air Blank		0.000	06:19
Subject Sample	# 1	0.105	06:20
Breath Volume		1.343	06:20
Air Blank		0.000	06:21
Air Blank		0.000	06:22
Subject Sample	# 2	0.125	06:23
Breath Volume		1.269	06:23
Air Blank		0.000	06:23
Control Test		0.080	06:24
Air Blank		0.000	06:24
Diagnostics Check		OK	06:24

Data Download Date/Time: 10/08/2020 13:04

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: STERN,ADAM
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 09/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 22:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: RANCHEZ

DATE/TIME OF TEST: 10/05/2020 23:31:19

Diagnostics Check		OK	23:23
Air Blank		0.000	23:24
Control Test		0.079	23:24
Air Blank		0.000	23:25
Subject Sample	# 1	0.215	23:26
Breath Volume		2.500	23:26
Air Blank		0.000	23:26
Air Blank		0.000	23:28
Subject Sample	# 2	0.208	23:29
Breath Volume		2.730	23:29
Air Blank		0.000	23:30
Control Test		0.079	23:30
Air Blank		0.000	23:31
Diagnostics Check		OK	23:31

Data Download Date/Time: 10/08/2020 13:04

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: STUEHRENBERG,ERIK
ARRESTING OFFICER AGENCY: Davie PD
LAST AGENCY INSPECTION: 09/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: BADILLO

DATE/TIME OF TEST: 10/06/2020 03:52:10

Diagnostics Check		OK	03:46	
Air Blank		0.000	03:46	
Control Test		0.080	03:46	
Air Blank		0.000	03:47	
Subject Sample	# 1	0.119	03:50	
Breath Volume		1.707	03:50	
Air Blank			03:51	Purge Fail
Air Blank			03:52	Purge Fail

Data Download Date/Time: 10/08/2020 13:04

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: STUEHRENBERG,ERIK
ARRESTING OFFICER AGENCY: Davie PD
LAST AGENCY INSPECTION: 09/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: BADILLO

DATE/TIME OF TEST: 10/06/2020 03:56:40

Diagnostics Check	OK	03:55	
Air Blank	0.000	03:55	
Control Test	0.049	03:56	Control Outside Tolerance
Air Blank	0.000	03:56	

Data Download Date/Time: 10/08/2020 13:04

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006926

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,JEFFREY,A
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: THOMPSON,JEFFREY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 10/20/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:58

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 10/27/2020 21:25:22

Diagnostics Check		OK	21:18
Air Blank		0.000	21:19
Control Test		0.080	21:19
Air Blank		0.000	21:20
Subject Sample	# 1	0.242	21:20
Breath Volume		2.410	21:20
Air Blank		0.000	21:21
Air Blank		0.000	21:23
Subject Sample	# 2	0.233	21:23
Breath Volume		1.402	21:23
Air Blank		0.000	21:24
Control Test		0.080	21:24
Air Blank		0.000	21:25
Diagnostics Check		OK	21:25

Data Download Date/Time: 11/30/2020 15:12

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006926

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,GARY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 10/20/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: COOKS

DATE/TIME OF TEST: 10/27/2020 23:42:51

Diagnostics Check		OK	23:35
Air Blank		0.000	23:36
Control Test		0.081	23:36
Air Blank		0.000	23:36
Subject Sample	# 1	0.064	23:37
Breath Volume		1.132	23:37
Air Blank		0.000	23:37
Air Blank		0.000	23:39
Subject Sample	# 2	0.065	23:41
Breath Volume		1.394	23:41
Air Blank		0.000	23:41
Control Test		0.084	23:42
Air Blank		0.000	23:42
Diagnostics Check		OK	23:42

Data Download Date/Time: 11/30/2020 15:12

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006937

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER,MONTREZ,R
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: YOUNG,THOMAS
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 09/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG923908
DRY GAS EXPIRY DATE: 08/27/2021
VIOLATION CODE: DUI
ARREST TIME: 05:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:52

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 10/12/2020 06:13:40

Diagnostics Check	OK	06:12	
Air Blank		06:13	RFI Detect
Air Blank	0.000	06:13	

Data Download Date/Time: 10/30/2020 11:48

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006937

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER,MONTREZ,R
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: YOUNG,THOMAS
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 09/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG923908
DRY GAS EXPIRY DATE: 08/27/2021
VIOLATION CODE: DUI
ARREST TIME: 05:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:52

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 10/12/2020 06:22:16

Diagnostics Check		OK	06:15
Air Blank		0.000	06:16
Control Test		0.079	06:16
Air Blank		0.000	06:16
Subject Sample	# 1	0.124	06:17
Breath Volume		1.441	06:17
Air Blank		0.000	06:18
Air Blank		0.000	06:20
Subject Sample	# 2	0.131	06:20
Breath Volume		1.609	06:20
Air Blank		0.000	06:21
Control Test		0.081	06:21
Air Blank		0.000	06:22
Diagnostics Check		OK	06:22

Data Download Date/Time: 10/30/2020 11:48

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006937

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER,MONTREZ,R
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: HUTCHINSON,JOSEPH
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 09/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG923908
DRY GAS EXPIRY DATE: 08/27/2021
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: LAND

DATE/TIME OF TEST: 10/17/2020 04:36:30

Diagnostics Check		OK	04:33	
Air Blank		0.000	04:34	
Control Test		0.082	04:34	
Air Blank		0.000	04:34	
Subject Sample	# 1		04:35	Subject Test Refused
Breath Volume		0.000	04:35	
Air Blank		0.000	04:35	
Control Test		0.082	04:35	
Air Blank		0.000	04:36	
Diagnostics Check		OK	04:36	

Data Download Date/Time: 10/30/2020 11:48

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006939

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEATING,KAYLA,C
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: KEATING,KAYLA
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 10/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: MILOPOULOS

DATE/TIME OF TEST: 10/30/2020 21:52:03

Diagnostics Check		OK	21:44
Air Blank		0.000	21:45
Control Test		0.076	21:45
Air Blank		0.000	21:46
Subject Sample	# 1	0.127	21:47
Breath Volume		3.300	21:47
Air Blank		0.000	21:48
Air Blank		0.000	21:49
Subject Sample	# 2	0.121	21:50
Breath Volume		3.433	21:50
Air Blank		0.000	21:51
Control Test		0.076	21:51
Air Blank		0.000	21:51
Diagnostics Check		OK	21:52

Data Download Date/Time: 11/18/2020 11:01

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007007

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STONER,NATHAN,C
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: STONER,NATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 10/08/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: DUNAWAY

DATE/TIME OF TEST: 10/24/2020 00:05:17

Diagnostics Check		OK	00:01	
Air Blank		0.000	00:01	
Control Test		0.077	00:02	
Air Blank		0.000	00:02	
Subject Sample	# 1		00:03	Subject Test Refused
Breath Volume		0.000	00:03	
Air Blank		0.000	00:04	
Control Test		0.078	00:04	
Air Blank		0.000	00:05	
Diagnostics Check		OK	00:05	

Data Download Date/Time: 11/23/2020 14:19

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007007

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STONER,NATHAN,C
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 10/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: GRADY

DATE/TIME OF TEST: 10/27/2020 23:00:50

Diagnostics Check		OK	22:57	
Air Blank		0.000	22:57	
Control Test		0.076	22:58	
Air Blank		0.000	22:58	
Subject Sample	# 1		22:59	Subject Test Refused
Breath Volume		0.000	22:59	
Air Blank		0.000	22:59	
Control Test		0.077	23:00	
Air Blank		0.000	23:00	
Diagnostics Check		OK	23:00	

Data Download Date/Time: 11/23/2020 14:19

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARMSTRONG,BRITTANY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ARMSTRONG,BRITTANY
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 09/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 0091908A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: GUZMAN

DATE/TIME OF TEST: 10/01/2020 02:03:40

Diagnostics Check		OK	01:57
Air Blank		0.000	01:57
Control Test		0.079	01:57
Air Blank		0.000	01:58
Subject Sample	# 1	0.157	01:58
Breath Volume		1.910	01:58
Air Blank		0.000	01:59
Air Blank		0.000	02:01
Subject Sample	# 2	0.162	02:01
Breath Volume		2.375	02:01
Air Blank		0.000	02:02
Control Test		0.077	02:03
Air Blank		0.000	02:03
Diagnostics Check		OK	02:03

Data Download Date/Time: 10/08/2020 10:58

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARMSTRONG,BRITTANY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ARMSTRONG,BRITTANY
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 09/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 0091908A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: RIGG

DATE/TIME OF TEST: 10/03/2020 23:30:03

Diagnostics Check		OK	23:23
Air Blank		0.000	23:24
Control Test		0.077	23:24
Air Blank		0.000	23:25
Subject Sample	# 1	0.073	23:25
Breath Volume		1.406	23:25
Air Blank		0.000	23:26
Air Blank		0.000	23:27
Subject Sample	# 2	0.066	23:28
Breath Volume		1.562	23:28
Air Blank		0.000	23:29
Control Test		0.077	23:29
Air Blank		0.000	23:29
Diagnostics Check		OK	23:30

Data Download Date/Time: 10/08/2020 10:58

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARMSTRONG,BRITTANY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 09/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 0091908A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: MENDOZA

DATE/TIME OF TEST: 10/04/2020 02:37:33

Diagnostics Check		OK	02:31
Air Blank		0.000	02:31
Control Test		0.077	02:31
Air Blank		0.000	02:32
Subject Sample	# 1	0.190	02:32
Breath Volume		3.031	02:32
Air Blank		0.000	02:33
Air Blank		0.000	02:35
Subject Sample	# 2	0.191	02:35
Breath Volume		1.304	02:35
Air Blank		0.000	02:36
Control Test		0.077	02:36
Air Blank		0.000	02:37
Diagnostics Check		OK	02:37

Data Download Date/Time: 10/08/2020 10:58

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARMSTRONG,BRITTANY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ARMSTRONG,BRITTANY
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 0091908A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: RIOFRIO

DATE/TIME OF TEST: 10/10/2020 01:11:22

Diagnostics Check		OK	01:04
Air Blank		0.000	01:05
Control Test		0.077	01:05
Air Blank		0.000	01:06
Subject Sample	# 1	0.151	01:06
Breath Volume		1.316	01:06
Air Blank		0.000	01:07
Air Blank		0.000	01:09
Subject Sample	# 2	0.154	01:09
Breath Volume		2.218	01:09
Air Blank		0.000	01:10
Control Test		0.078	01:10
Air Blank		0.000	01:11
Diagnostics Check		OK	01:11

Data Download Date/Time: 11/23/2020 14:22

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARMSTRONG,BRITTANY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LIPTAK,LAUREN
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 10/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 0091908A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: SPIRES

DATE/TIME OF TEST: 10/23/2020 04:44:24

Diagnostics Check		OK	04:36
Air Blank		0.000	04:36
Control Test		0.077	04:37
Air Blank		0.000	04:37
Subject Sample	# 1	0.123	04:39
Breath Volume		1.195	04:39
Air Blank		0.000	04:40
Air Blank		0.000	04:42
Subject Sample	# 2	0.119	04:42
Breath Volume		1.289	04:42
Air Blank		0.000	04:43
Control Test		0.076	04:43
Air Blank		0.000	04:44
Diagnostics Check		OK	04:44

Data Download Date/Time: 11/23/2020 14:23

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARMSTRONG,BRITTANY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ARMSTRONG,BRITTANY
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 0091908A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:16

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 10/28/2020 22:56:11

Diagnostics Check		OK	22:44	
Air Blank		0.000	22:44	
Control Test		0.076	22:45	
Air Blank		0.000	22:45	
Subject Sample	# 1	0.118	22:48	Volume Not Met
Breath Volume		0.082	22:48	
Air Blank		0.000	22:49	
Air Blank		0.000	22:51	
Subject Sample	# 2	0.088	22:54	Volume Not Met
Breath Volume		0.031	22:54	
Air Blank		0.000	22:55	
Control Test		0.077	22:55	
Air Blank		0.000	22:56	
Diagnostics Check		OK	22:56	

Data Download Date/Time: 11/23/2020 14:23

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARMSTRONG,BRITTANY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ARMSTRONG,BRITTANY
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 0091908A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 10/31/2020 03:38:44

Diagnostics Check		OK	03:28
Air Blank		0.000	03:29
Control Test		0.077	03:29
Air Blank		0.000	03:30
Subject Sample	# 1	0.162	03:32
Breath Volume		1.820	03:32
Air Blank		0.000	03:33
Air Blank		0.000	03:35
Subject Sample	# 2	0.160	03:36
Breath Volume		1.562	03:36
Air Blank		0.000	03:37
Control Test		0.077	03:38
Air Blank		0.000	03:38
Diagnostics Check		OK	03:38

Data Download Date/Time: 11/23/2020 14:23

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007009

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JAMES,W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: STONER,NATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 09/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: AYALA SARAVIA

DATE/TIME OF TEST: 10/04/2020 00:59:35

Diagnostics Check		OK	00:50
Air Blank		0.000	00:50
Control Test		0.079	00:51
Air Blank		0.000	00:51
Subject Sample	# 1	0.291	00:54
Breath Volume		1.644	00:54
Air Blank		0.000	00:55
Air Blank		0.000	00:56
Subject Sample	# 2	0.274	00:57
Breath Volume		2.742	00:57
Air Blank		0.000	00:58
Control Test		0.077	00:58
Air Blank		0.000	00:59
Diagnostics Check		OK	00:59

Data Download Date/Time: 10/08/2020 11:08

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007009

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STONER,NATHAN,C
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 09/17/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:04

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 10/04/2020 21:33:19

Diagnostics Check		OK	21:24	
Air Blank		0.000	21:25	
Control Test		0.079	21:25	
Air Blank		0.000	21:26	
Subject Sample	# 1	0.195	21:29	Volume Not Met
Breath Volume		0.550	21:29	
Air Blank		0.000	21:30	
Air Blank		0.000	21:31	
Subject Sample	# 2		21:31	Subject Test Refused
Breath Volume		0.000	21:31	
Air Blank		0.000	21:32	
Control Test		0.077	21:32	
Air Blank		0.000	21:33	
Diagnostics Check		OK	21:33	

Data Download Date/Time: 10/08/2020 11:08

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007009

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STONER,NATHAN,C
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 10/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: RUIZ PORTILLA

DATE/TIME OF TEST: 10/13/2020 00:05:57

Diagnostics Check		OK	23:59
Air Blank		0.000	23:59
Control Test		0.078	23:59
Air Blank		0.000	00:00
Subject Sample	# 1	0.000	00:01
Breath Volume		1.242	00:01
Air Blank		0.000	00:01
Air Blank		0.000	00:03
Subject Sample	# 2	0.000	00:04
Breath Volume		1.667	00:04
Air Blank		0.000	00:05
Control Test		0.079	00:05
Air Blank		0.000	00:05
Diagnostics Check		OK	00:05

Data Download Date/Time: 11/18/2020 12:49

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007009

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JAMES,W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: STONER,NATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 10/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: DUMERVIL

DATE/TIME OF TEST: 10/18/2020 00:19:56

Diagnostics Check		OK	00:13
Air Blank		0.000	00:13
Control Test		0.077	00:14
Air Blank		0.000	00:14
Subject Sample	# 1	0.129	00:15
Breath Volume		3.867	00:15
Air Blank		0.000	00:16
Air Blank		0.000	00:17
Subject Sample	# 2	0.127	00:18
Breath Volume		4.328	00:18
Air Blank		0.000	00:19
Control Test		0.079	00:19
Air Blank		0.000	00:19
Diagnostics Check		OK	00:19

Data Download Date/Time: 11/18/2020 12:49

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKS,D,M
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: PARKS,D
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 10/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: MIHAILOV

DATE/TIME OF TEST: 10/31/2020 19:55:00

Diagnostics Check		OK	19:48
Air Blank		0.000	19:49
Control Test		0.080	19:49
Air Blank		0.000	19:49
Subject Sample	# 1	0.315	19:50
Breath Volume		2.687	19:50
Air Blank		0.000	19:51
Air Blank		0.000	19:52
Subject Sample	# 2	0.323	19:53
Breath Volume		1.308	19:53
Air Blank		0.000	19:54
Control Test		0.079	19:54
Air Blank		0.000	19:54
Diagnostics Check		OK	19:55

Data Download Date/Time: 11/10/2020 15:10

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIBBONS,S,L
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: CAREY,E
ARRESTING OFFICER AGENCY: MARCO ISLAND PD
LAST AGENCY INSPECTION: 10/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: BURKE

DATE/TIME OF TEST: 10/31/2020 23:46:10

Diagnostics Check		OK	23:37
Air Blank		0.000	23:37
Control Test		0.080	23:37
Air Blank		0.000	23:38
Subject Sample	# 1	0.189	23:39
Breath Volume		1.468	23:39
Air Blank		0.000	23:39
Air Blank		0.000	23:41
Subject Sample	# 2	0.196	23:44
Breath Volume		2.226	23:44
Air Blank		0.000	23:45
Control Test		0.080	23:45
Air Blank		0.000	23:46
Diagnostics Check		OK	23:46

Data Download Date/Time: 11/10/2020 15:10

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTIERREZ,KRISTINA,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: GUTIERREZ,KRISTINA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: FERRUFINO

DATE/TIME OF TEST: 10/01/2020 02:00:09

Diagnostics Check		OK	01:57	
Air Blank		0.000	01:57	
Control Test		0.079	01:57	
Air Blank		0.000	01:58	
Subject Sample	# 1		01:58	Subject Test Refused
Breath Volume		0.000	01:58	
Air Blank		0.000	01:59	
Control Test		0.080	01:59	
Air Blank		0.000	02:00	
Diagnostics Check		OK	02:00	

Data Download Date/Time: 10/28/2020 11:28

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTIERREZ,KRISTINA,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: GUTIERREZ,KRISTINA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: RODRIGUEZ PADRON

DATE/TIME OF TEST: 10/02/2020 02:16:37

Diagnostics Check		OK	02:13	
Air Blank		0.000	02:14	
Control Test		0.080	02:14	
Air Blank		0.000	02:15	
Subject Sample	# 1		02:15	Subject Test Refused
Breath Volume		0.000	02:15	
Air Blank		0.000	02:15	
Control Test		0.080	02:15	
Air Blank		0.000	02:16	
Diagnostics Check		OK	02:16	

Data Download Date/Time: 10/28/2020 11:28

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLORIAN,ROLANDO,I
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FLORIAN,ROLANDO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 10/02/2020 22:07:17

Diagnostics Check		OK	21:58	
Air Blank		0.000	21:58	
Control Test		0.080	21:59	
Air Blank		0.000	21:59	
Subject Sample	# 1	0.132	22:02	Volume Not Met
Breath Volume		0.234	22:02	
Air Blank		0.000	22:03	
Air Blank		0.000	22:05	
Subject Sample	# 2		22:05	Subject Test Refused
Breath Volume		0.000	22:05	
Air Blank		0.000	22:06	
Control Test		0.077	22:06	
Air Blank		0.000	22:07	
Diagnostics Check		OK	22:07	

Data Download Date/Time: 10/28/2020 11:28

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHALA,DANIEL,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: GUTIERREZ,JORGE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 10/03/2020 03:04:44

Diagnostics Check		OK	02:58
Air Blank		0.000	02:58
Control Test		0.081	02:58
Air Blank		0.000	02:59
Subject Sample	# 1	0.201	02:59
Breath Volume		2.582	02:59
Air Blank		0.000	03:00
Air Blank		0.000	03:02
Subject Sample	# 2	0.193	03:03
Breath Volume		2.503	03:03
Air Blank		0.000	03:03
Control Test		0.077	03:04
Air Blank		0.000	03:04
Diagnostics Check		OK	03:04

Data Download Date/Time: 10/28/2020 11:28

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLORIAN,ROLANDO,I
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FLORIAN,ROLANDO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: FERRER

DATE/TIME OF TEST: 10/11/2020 22:20:36

Diagnostics Check		OK	22:18	
Air Blank		0.000	22:18	
Control Test		0.080	22:19	
Air Blank		0.000	22:19	
Subject Sample	# 1		22:19	Subject Test Refused
Breath Volume		0.000	22:19	
Air Blank			22:20	RFI Detect
Air Blank		0.000	22:20	

Data Download Date/Time: 10/28/2020 11:28

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAVINON,CARLOS,M
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: SAVINON,CARLOS
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: CUROTTO

DATE/TIME OF TEST: 10/12/2020 04:08:03

Diagnostics Check		OK	03:56	
Air Blank		0.000	03:56	
Control Test		0.080	03:57	
Air Blank		0.000	03:57	
Subject Sample	# 1	0.130	04:00	Volume Not Met
Breath Volume		0.625	04:00	
Air Blank		0.000	04:01	
Air Blank		0.000	04:03	
Subject Sample	# 2	0.157	04:06	Volume Not Met
Breath Volume		0.957	04:06	
Air Blank		0.000	04:07	
Control Test		0.079	04:07	
Air Blank		0.000	04:07	
Diagnostics Check		OK	04:08	

Data Download Date/Time: 10/28/2020 11:28

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAVINON,CARLOS,M
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: SAVINON,CARLOS
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: CUROTTO

DATE/TIME OF TEST: 10/12/2020 04:57:01

Diagnostics Check		OK	04:43	
Air Blank		0.000	04:44	
Control Test		0.079	04:44	
Air Blank		0.000	04:44	
Subject Sample	# 1	0.117	04:46	
Breath Volume		1.437	04:46	
Air Blank		0.000	04:47	
Air Blank		0.000	04:48	
Subject Sample	# 2	0.121	04:51	Volume Not Met
Breath Volume		0.437	04:51	
Air Blank		0.000	04:52	
Air Blank		0.000	04:54	
Subject Sample	# 3	0.116	04:55	
Breath Volume		1.296	04:55	
Air Blank		0.000	04:56	
Control Test		0.077	04:56	
Air Blank		0.000	04:56	
Diagnostics Check		OK	04:57	

Data Download Date/Time: 10/28/2020 11:28

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAVINON,CARLOS,M
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: SAVINON,CARLOS
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: ARANGO

DATE/TIME OF TEST: 10/20/2020 03:08:32

Diagnostics Check		OK	03:02
Air Blank		0.000	03:02
Control Test		0.080	03:03
Air Blank		0.000	03:03
Subject Sample	# 1	0.122	03:04
Breath Volume		2.144	03:04
Air Blank		0.000	03:04
Air Blank		0.000	03:06
Subject Sample	# 2	0.114	03:06
Breath Volume		2.363	03:06
Air Blank		0.000	03:07
Control Test		0.080	03:07
Air Blank		0.000	03:08
Diagnostics Check		OK	03:08

Data Download Date/Time: 10/28/2020 11:28

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUAREZ,CARLOS,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: LORENTE,GIOVANNI
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: CALDERON-FERNANDEZ

DATE/TIME OF TEST: 10/25/2020 01:20:13

Diagnostics Check		OK	01:11
Air Blank		0.000	01:11
Control Test		0.080	01:12
Air Blank		0.000	01:12
Subject Sample	# 1	0.267	01:14
Breath Volume		2.070	01:14
Air Blank		0.000	01:15
Air Blank		0.000	01:17
Subject Sample	# 2	0.262	01:18
Breath Volume		1.636	01:18
Air Blank		0.000	01:19
Control Test		0.079	01:19
Air Blank		0.000	01:20
Diagnostics Check		OK	01:20

Data Download Date/Time: 10/28/2020 11:28

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNE,STEVEN,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: THORNE,STEVEN
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: ACEVEDO LOPEZ

DATE/TIME OF TEST: 10/24/2020 02:41:50

Diagnostics Check		OK	02:35
Air Blank		0.000	02:35
Control Test		0.078	02:36
Air Blank		0.000	02:36
Subject Sample	# 1	0.165	02:37
Breath Volume		2.828	02:37
Air Blank		0.000	02:37
Air Blank		0.000	02:39
Subject Sample	# 2	0.166	02:40
Breath Volume		2.886	02:40
Air Blank		0.000	02:40
Control Test		0.077	02:41
Air Blank		0.000	02:41
Diagnostics Check		OK	02:41

Data Download Date/Time: 11/18/2020 12:40

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNE,S,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 10/25/2020 01:31:26

Diagnostics Check		OK	01:25	
Air Blank		0.000	01:26	
Control Test		0.078	01:26	
Air Blank		0.000	01:26	
Subject Sample	# 1	0.149	01:27	
Breath Volume		3.539	01:27	
Air Blank		0.000	01:28	
Air Blank		0.000	01:29	
Subject Sample	# 2		01:30	RFI Detect
Breath Volume		0.000	01:30	
Air Blank		0.000	01:30	
Control Test		0.078	01:30	
Air Blank		0.000	01:31	
Diagnostics Check		OK	01:31	

Data Download Date/Time: 11/18/2020 12:40

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNE,S,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 10/25/2020 01:40:25

Diagnostics Check		OK	01:33
Air Blank		0.000	01:34
Control Test		0.078	01:34
Air Blank		0.000	01:35
Subject Sample	# 1	0.137	01:35
Breath Volume		3.445	01:35
Air Blank		0.000	01:36
Air Blank		0.000	01:38
Subject Sample	# 2	0.132	01:38
Breath Volume		2.265	01:38
Air Blank		0.000	01:39
Control Test		0.078	01:39
Air Blank		0.000	01:40
Diagnostics Check		OK	01:40

Data Download Date/Time: 11/18/2020 12:40

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNE,STEVEN,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FISHER,THOMAS
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: VILSAINT

DATE/TIME OF TEST: 10/25/2020 03:58:52

Diagnostics Check		OK	03:49
Air Blank		0.000	03:49
Control Test		0.078	03:50
Air Blank		0.000	03:50
Subject Sample	# 1	0.112	03:52
Breath Volume		1.421	03:52
Air Blank		0.000	03:53
Air Blank		0.000	03:55
Subject Sample	# 2	0.111	03:57
Breath Volume		1.292	03:57
Air Blank		0.000	03:57
Control Test		0.077	03:58
Air Blank		0.000	03:58
Diagnostics Check		OK	03:58

Data Download Date/Time: 11/18/2020 12:40

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,BRENDAN,D
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: BROWN,BRENDAN
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 10/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1160200
DRY GAS EXPIRY DATE: 06/20/2022
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:19

SUBJECT'S LAST NAME: BELK

DATE/TIME OF TEST: 10/30/2020 21:50:00

Diagnostics Check		OK	21:39
Air Blank		0.000	21:40
Control Test		0.082	21:40
Air Blank		0.000	21:40
Subject Sample	# 1	0.241	21:43
Breath Volume		2.492	21:43
Air Blank		0.000	21:44
Air Blank		0.000	21:45
Subject Sample	# 2	0.234	21:48
Breath Volume		2.328	21:48
Air Blank		0.000	21:49
Control Test		0.081	21:49
Air Blank		0.000	21:49
Diagnostics Check		OK	21:50

Data Download Date/Time: 11/09/2020 13:47

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,BRENDAN,D
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: BROWN,BRENDAN
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 10/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1160200
DRY GAS EXPIRY DATE: 06/20/2022
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: LOTT

DATE/TIME OF TEST: 10/31/2020 03:24:26

Diagnostics Check		OK	03:16
Air Blank		0.000	03:16
Control Test		0.082	03:16
Air Blank		0.000	03:17
Subject Sample	# 1	0.168	03:18
Breath Volume		1.175	03:18
Air Blank		0.000	03:19
Air Blank		0.000	03:21
Subject Sample	# 2	0.178	03:22
Breath Volume		2.675	03:22
Air Blank		0.000	03:23
Control Test		0.082	03:23
Air Blank		0.000	03:24
Diagnostics Check		OK	03:24

Data Download Date/Time: 11/09/2020 13:47

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOBSON,BEAU,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: TORRES CASTRO

DATE/TIME OF TEST: 10/03/2020 02:20:58

Diagnostics Check		OK	02:13
Air Blank		0.000	02:14
Control Test		0.079	02:14
Air Blank		0.000	02:14
Subject Sample	# 1	0.121	02:15
Breath Volume		1.496	02:15
Air Blank		0.000	02:16
Air Blank		0.000	02:18
Subject Sample	# 2	0.130	02:19
Breath Volume		3.523	02:19
Air Blank		0.000	02:20
Control Test		0.078	02:20
Air Blank		0.000	02:20
Diagnostics Check		OK	02:20

Data Download Date/Time: 10/20/2020 09:44

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWOGGER,EDWARD,A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: SWOGGER,EDWARD
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 09/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:03

SUBJECT'S LAST NAME: RUTLEDGE

DATE/TIME OF TEST: 10/02/2020 20:27:59

Diagnostics Check		OK	20:23	
Air Blank		0.000	20:24	
Control Test		0.080	20:24	
Air Blank		0.000	20:25	
Subject Sample	# 1		20:26	Subject Test Refused
Breath Volume		0.000	20:26	
Air Blank		0.000	20:27	
Control Test		0.079	20:27	
Air Blank		0.000	20:27	
Diagnostics Check		OK	20:27	

Data Download Date/Time: 10/06/2020 11:12

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TELLIER,JEREMEY,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: BOZZETTI,AVERY
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 09/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 10/03/2020 04:20:11

Diagnostics Check		OK	04:13
Air Blank		0.000	04:13
Control Test		0.079	04:13
Air Blank		0.000	04:14
Subject Sample	# 1	0.227	04:15
Breath Volume		1.558	04:15
Air Blank		0.000	04:16
Air Blank		0.000	04:17
Subject Sample	# 2	0.235	04:18
Breath Volume		2.347	04:18
Air Blank		0.000	04:19
Control Test		0.077	04:19
Air Blank		0.000	04:20
Diagnostics Check		OK	04:20

Data Download Date/Time: 10/06/2020 11:12

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOGERLEY,NEIL,F
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: WAITE,KENNITH
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 09/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:13

SUBJECT'S LAST NAME: HAYEES

DATE/TIME OF TEST: 10/03/2020 19:37:26

Diagnostics Check		OK	19:34	
Air Blank		0.000	19:35	
Control Test		0.080	19:35	
Air Blank		0.000	19:35	
Subject Sample	# 1		19:36	Subject Test Refused
Breath Volume		0.000	19:36	
Air Blank		0.000	19:36	
Control Test		0.079	19:36	
Air Blank		0.000	19:37	
Diagnostics Check		OK	19:37	

Data Download Date/Time: 10/06/2020 11:12

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TELLIER,JEREMEY,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: TELLIER,JEREMEY
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 10/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1217042
DRY GAS EXPIRY DATE: 10/10/2022
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: HAVEL

DATE/TIME OF TEST: 10/09/2020 23:47:06

Diagnostics Check		OK	23:40
Air Blank		0.000	23:41
Control Test		0.080	23:41
Air Blank		0.000	23:42
Subject Sample	# 1	0.216	23:42
Breath Volume		3.363	23:42
Air Blank		0.000	23:43
Air Blank		0.000	23:45
Subject Sample	# 2	0.196	23:45
Breath Volume		2.222	23:45
Air Blank		0.000	23:46
Control Test		0.077	23:46
Air Blank		0.000	23:47
Diagnostics Check		OK	23:47

Data Download Date/Time: 11/18/2020 11:55

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUDSON,ANDREW,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: HUDSON,ANDREW
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 10/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1217042
DRY GAS EXPIRY DATE: 10/10/2022
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: TRIPP

DATE/TIME OF TEST: 10/11/2020 00:13:43

Diagnostics Check		OK	00:10	
Air Blank		0.000	00:11	
Control Test		0.081	00:11	
Air Blank		0.000	00:12	
Subject Sample	# 1		00:12	Subject Test Refused
Breath Volume		0.000	00:12	
Air Blank		0.000	00:12	
Control Test		0.080	00:13	
Air Blank		0.000	00:13	
Diagnostics Check		OK	00:13	

Data Download Date/Time: 11/18/2020 11:55

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TELLIER,JEREMEY,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: TELLIER,JEREMEY
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 10/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1217042
DRY GAS EXPIRY DATE: 10/10/2022
VIOLATION CODE: DUI
ARREST TIME: 20:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:32

SUBJECT'S LAST NAME: MULLARKEY

DATE/TIME OF TEST: 10/14/2020 20:57:49

Diagnostics Check		OK	20:53	
Air Blank		0.000	20:53	
Control Test		0.080	20:54	
Air Blank		0.000	20:54	
Subject Sample	# 1		20:56	Subject Test Refused
Breath Volume		0.000	20:56	
Air Blank		0.000	20:56	
Control Test		0.080	20:57	
Air Blank		0.000	20:57	
Diagnostics Check		OK	20:57	

Data Download Date/Time: 11/18/2020 11:55

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUDSON,ANDREW,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: HUDSON,ANDREW
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 10/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1217042
DRY GAS EXPIRY DATE: 10/10/2022
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: ARROYAVE

DATE/TIME OF TEST: 10/16/2020 02:13:03

Diagnostics Check		OK	02:09	
Air Blank		0.000	02:10	
Control Test		0.081	02:10	
Air Blank		0.000	02:11	
Subject Sample	# 1		02:11	Subject Test Refused
Breath Volume		0.000	02:11	
Air Blank		0.000	02:12	
Control Test		0.081	02:12	
Air Blank		0.000	02:12	
Diagnostics Check		OK	02:13	

Data Download Date/Time: 11/18/2020 11:55

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUDSON,ANDREW,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: HUDSON,ANDREW
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 10/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1217042
DRY GAS EXPIRY DATE: 10/10/2022
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: TORRES JR

DATE/TIME OF TEST: 10/20/2020 02:13:11

Diagnostics Check		OK	02:06
Air Blank		0.000	02:07
Control Test		0.080	02:07
Air Blank		0.000	02:07
Subject Sample	# 1	0.160	02:08
Breath Volume		1.468	02:08
Air Blank		0.000	02:09
Air Blank		0.000	02:10
Subject Sample	# 2	0.168	02:11
Breath Volume		2.171	02:11
Air Blank		0.000	02:12
Control Test		0.080	02:12
Air Blank		0.000	02:13
Diagnostics Check		OK	02:13

Data Download Date/Time: 11/18/2020 11:55

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TELLIER,JEREMEY,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: HUDSON,ANDREW
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 10/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1217042
DRY GAS EXPIRY DATE: 10/10/2022
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: CASTLE

DATE/TIME OF TEST: 10/21/2020 04:27:11

Diagnostics Check		OK	04:20
Air Blank		0.000	04:21
Control Test		0.080	04:21
Air Blank		0.000	04:22
Subject Sample	# 1	0.233	04:22
Breath Volume		1.265	04:22
Air Blank		0.000	04:23
Air Blank		0.000	04:25
Subject Sample	# 2	0.240	04:25
Breath Volume		1.105	04:25
Air Blank		0.000	04:26
Control Test		0.079	04:26
Air Blank		0.000	04:27
Diagnostics Check		OK	04:27

Data Download Date/Time: 11/18/2020 11:55

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWOGGER,EDWARD,A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: SWOGGER,EDWARD
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 10/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1217042
DRY GAS EXPIRY DATE: 10/10/2022
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: RICCI

DATE/TIME OF TEST: 10/22/2020 00:53:47

Diagnostics Check		OK	00:50	
Air Blank		0.000	00:51	
Control Test		0.080	00:51	
Air Blank		0.000	00:52	
Subject Sample	# 1		00:52	Subject Test Refused
Breath Volume		0.000	00:52	
Air Blank		0.000	00:52	
Control Test		0.080	00:53	
Air Blank		0.000	00:53	
Diagnostics Check		OK	00:53	

Data Download Date/Time: 11/18/2020 11:55

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUDSON,ANDREW,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: HUDSON,ANDREW
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 10/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1217042
DRY GAS EXPIRY DATE: 10/10/2022
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: DENISAC

DATE/TIME OF TEST: 10/24/2020 00:52:00

Diagnostics Check		OK	00:46	
Air Blank		0.000	00:47	
Control Test		0.080	00:47	
Air Blank		0.000	00:48	
Subject Sample	# 1		00:50	Subject Test Refused
Breath Volume		0.000	00:50	
Air Blank		0.000	00:51	
Control Test		0.081	00:51	
Air Blank		0.000	00:51	
Diagnostics Check		OK	00:52	

Data Download Date/Time: 11/18/2020 11:55

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWOGGER,EDWARD,A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MESSIER,MARTIN
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 10/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1217042
DRY GAS EXPIRY DATE: 10/10/2022
VIOLATION CODE: DUI
ARREST TIME: 18:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: DALENBERG

DATE/TIME OF TEST: 10/24/2020 19:26:15

Diagnostics Check		OK	19:23	
Air Blank		0.000	19:23	
Control Test		0.080	19:23	
Air Blank		0.000	19:24	
Subject Sample	# 1		19:24	Subject Test Refused
Breath Volume		0.000	19:24	
Air Blank		0.000	19:25	
Control Test		0.080	19:25	
Air Blank		0.000	19:26	
Diagnostics Check		OK	19:26	

Data Download Date/Time: 11/18/2020 11:55

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODDY,AARON,A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: GILROY,ZACHARY
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 10/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1217042
DRY GAS EXPIRY DATE: 10/10/2022
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: HARCUP

DATE/TIME OF TEST: 10/31/2020 00:27:11

Diagnostics Check		OK	00:20
Air Blank		0.000	00:20
Control Test		0.080	00:21
Air Blank		0.000	00:21
Subject Sample	# 1	0.195	00:22
Breath Volume		1.621	00:22
Air Blank		0.000	00:23
Air Blank		0.000	00:24
Subject Sample	# 2	0.203	00:25
Breath Volume		1.234	00:25
Air Blank		0.000	00:26
Control Test		0.080	00:26
Air Blank		0.000	00:27
Diagnostics Check		OK	00:27

Data Download Date/Time: 11/18/2020 11:55

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007169

INSTRUMENT REGISTERED TO: FWC TAMPA
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETTIFER,JAMES,P
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: PASS,JUSTIN
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09820080A2
DRY GAS EXPIRY DATE: 06/05/2022
VIOLATION CODE: BUI
ARREST TIME: 16:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 10/17/2020 17:22:41

Diagnostics Check	OK	17:21	
Air Blank	0.000	17:21	
Control Test	0.000	17:22	Control Outside Tolerance
Air Blank	0.000	17:22	

Data Download Date/Time: 11/17/2020 14:01

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007169

INSTRUMENT REGISTERED TO: FWC TAMPA
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETTIFER,JAMES,P
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: PASS,JUSTIN
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09820080A2
DRY GAS EXPIRY DATE: 06/05/2022
VIOLATION CODE: BUI
ARREST TIME: 16:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 10/17/2020 17:26:35

Diagnostics Check	OK	17:25	
Air Blank	0.000	17:25	
Control Test	0.000	17:26	Control Outside Tolerance
Air Blank	0.000	17:26	

Data Download Date/Time: 11/17/2020 14:01

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007169

INSTRUMENT REGISTERED TO: FWC TAMPA
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETTIFER,JAMES,P
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: PASS,JUSTIN
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09820080A2
DRY GAS EXPIRY DATE: 06/05/2022
VIOLATION CODE: BUI
ARREST TIME: 16:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 10/17/2020 17:33:05

Diagnostics Check	OK	17:31	
Air Blank	0.000	17:32	
Control Test	0.000	17:32	Control Outside Tolerance
Air Blank	0.000	17:33	

Data Download Date/Time: 11/17/2020 14:01

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007169

INSTRUMENT REGISTERED TO: FWC TAMPA
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETTIFER,JAMES,P
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: PASS,JUSTIN
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09820080A2
DRY GAS EXPIRY DATE: 06/05/2022
VIOLATION CODE: BUI
ARREST TIME: 16:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 10/17/2020 17:42:19

Diagnostics Check	OK	17:40	
Air Blank	0.000	17:40	
Control Test	0.079	17:41	
Air Blank	0.000	17:41	
		17:41	RFI Detect
Air Blank	0.000	17:42	

Data Download Date/Time: 11/17/2020 14:01

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007169

INSTRUMENT REGISTERED TO: FWC TAMPA
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETTIFER,JAMES,P
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: PASS,JUSTIN
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09820080A2
DRY GAS EXPIRY DATE: 06/05/2022
VIOLATION CODE: BUI
ARREST TIME: 16:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 10/17/2020 17:48:37

Diagnostics Check		OK	17:45	
Air Blank		0.000	17:45	
Control Test		0.079	17:45	
Air Blank		0.000	17:46	
Subject Sample	# 1		17:46	Subject Test Refused
Breath Volume		0.000	17:46	
Air Blank		0.000	17:47	
Control Test		0.079	17:47	
Air Blank		0.000	17:48	
Diagnostics Check		Fail	17:48	Internal Printer Test
Air Blank		0.000	17:48	

Data Download Date/Time: 11/17/2020 14:01

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007169

INSTRUMENT REGISTERED TO: FWC TAMPA
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PULASKI,DAMON,J
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: PULASKI,DAMON
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 10/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09820080A2
DRY GAS EXPIRY DATE: 06/05/2022
VIOLATION CODE: BUI
ARREST TIME: 18:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:28

SUBJECT'S LAST NAME: QUILES

DATE/TIME OF TEST: 10/18/2020 18:59:18

Diagnostics Check		OK	18:50
Air Blank		0.000	18:50
Control Test		0.078	18:50
Air Blank		0.000	18:51
Subject Sample	# 1	0.096	18:53
Breath Volume		1.492	18:53
Air Blank		0.000	18:53
Air Blank		0.000	18:55
Subject Sample	# 2	0.097	18:57
Breath Volume		1.683	18:57
Air Blank		0.000	18:58
Control Test		0.078	18:58
Air Blank		0.000	18:59
Diagnostics Check		OK	18:59

Data Download Date/Time: 11/17/2020 14:01

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HRADECKY,CORBIN,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: BRAGG,GARRET
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 09/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1217042
DRY GAS EXPIRY DATE: 10/10/2022
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: MAKI

DATE/TIME OF TEST: 10/03/2020 21:43:36

Diagnostics Check		OK	21:40	
Air Blank		0.000	21:40	
Control Test		0.081	21:40	
Air Blank		0.000	21:41	
Subject Sample	# 1		21:42	Subject Test Refused
Breath Volume		0.000	21:42	
Air Blank		0.000	21:42	
Control Test		0.081	21:43	
Air Blank		0.000	21:43	
Diagnostics Check		OK	21:43	

Data Download Date/Time: 10/07/2020 10:56

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUERRA,WILLIAM,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: BRAGG,GARETT
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 09/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1217042
DRY GAS EXPIRY DATE: 10/10/2022
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: EBERT

DATE/TIME OF TEST: 10/04/2020 22:37:59

Diagnostics Check		OK	22:30
Air Blank		0.000	22:31
Control Test		0.081	22:31
Air Blank		0.000	22:31
Subject Sample	# 1	0.091	22:33
Breath Volume		2.621	22:33
Air Blank		0.000	22:34
Air Blank		0.000	22:35
Subject Sample	# 2	0.094	22:36
Breath Volume		2.539	22:36
Air Blank		0.000	22:37
Control Test		0.079	22:37
Air Blank		0.000	22:37
Diagnostics Check		OK	22:37

Data Download Date/Time: 10/07/2020 10:56

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HRADECKY,CORBIN,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HRADECKY,CORBIN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 09/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1217042
DRY GAS EXPIRY DATE: 10/10/2022
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 10/05/2020 00:50:37

Diagnostics Check		OK	00:43
Air Blank		0.000	00:44
Control Test		0.080	00:44
Air Blank		0.000	00:45
Subject Sample	# 1	0.179	00:45
Breath Volume		1.683	00:45
Air Blank		0.000	00:46
Air Blank		0.000	00:48
Subject Sample	# 2	0.176	00:48
Breath Volume		1.933	00:48
Air Blank		0.000	00:49
Control Test		0.079	00:50
Air Blank		0.000	00:50
Diagnostics Check		OK	00:50

Data Download Date/Time: 10/07/2020 10:56

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HRADECKY,CORBIN,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HRADECKY,CORBIN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 10/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1217042
DRY GAS EXPIRY DATE: 10/10/2022
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: PANKEY

DATE/TIME OF TEST: 10/11/2020 04:26:50

Diagnostics Check		OK	04:23	
Air Blank		0.000	04:23	
Control Test		0.080	04:23	
Air Blank		0.000	04:24	
Subject Sample	# 1		04:25	Subject Test Refused
Breath Volume		0.000	04:25	
Air Blank		0.000	04:25	
Control Test		0.080	04:26	
Air Blank		0.000	04:26	
Diagnostics Check		OK	04:26	

Data Download Date/Time: 11/19/2020 11:25

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HRADECKY,CORBIN,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HRADECKY,CORBIN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 10/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1217042
DRY GAS EXPIRY DATE: 10/10/2022
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: SIMCIC

DATE/TIME OF TEST: 10/11/2020 22:34:27

Diagnostics Check		OK	22:21	
Air Blank		0.000	22:22	
Control Test		0.081	22:22	
Air Blank		0.000	22:22	
Subject Sample	# 1	0.166	22:24	
Breath Volume		1.863	22:24	
Air Blank		0.000	22:24	
Air Blank		0.000	22:26	
Subject Sample	# 2		22:29	No Sample Provided
Breath Volume		0.000	22:29	
Air Blank		0.000	22:30	
Air Blank		0.000	22:32	
Subject Sample	# 3	0.162	22:32	
Breath Volume		1.582	22:32	
Air Blank		0.000	22:33	
Control Test		0.080	22:33	
Air Blank		0.000	22:34	
Diagnostics Check		OK	22:34	

Data Download Date/Time: 11/19/2020 11:25

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HRADECKY,CORBIN,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: WHITE,BRANDON
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 10/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1217042
DRY GAS EXPIRY DATE: 10/10/2022
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: HREHA

DATE/TIME OF TEST: 10/14/2020 00:11:26

Diagnostics Check		OK	00:05
Air Blank		0.000	00:05
Control Test		0.080	00:05
Air Blank		0.000	00:06
Subject Sample	# 1	0.199	00:06
Breath Volume		2.492	00:06
Air Blank		0.000	00:07
Air Blank		0.000	00:09
Subject Sample	# 2	0.214	00:09
Breath Volume		1.292	00:09
Air Blank		0.000	00:10
Control Test		0.080	00:10
Air Blank		0.000	00:11
Diagnostics Check		OK	00:11

Data Download Date/Time: 11/19/2020 11:25

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WESTERBAND,FRANK,M
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: RODRIGUEZ,JONATHAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 10/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1217042
DRY GAS EXPIRY DATE: 10/10/2022
VIOLATION CODE: DUI
ARREST TIME: 06:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:19

SUBJECT'S LAST NAME: VALIENTE

DATE/TIME OF TEST: 10/18/2020 07:51:23

Diagnostics Check		OK	07:39	
Air Blank		0.000	07:40	
Control Test		0.080	07:40	
Air Blank		0.000	07:41	
Subject Sample	# 1		07:44	No Sample Provided
Breath Volume		0.000	07:44	
Air Blank		0.000	07:44	
Air Blank		0.000	07:46	
Subject Sample	# 2		07:50	No Sample Provided
Breath Volume		0.000	07:50	
Air Blank		0.000	07:50	
Control Test		0.080	07:50	
Air Blank		0.000	07:51	
Diagnostics Check		OK	07:51	

Data Download Date/Time: 11/19/2020 11:25

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARRILLO,JULIO,C
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: MARINO,JUAN CARLOS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 10/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1217042
DRY GAS EXPIRY DATE: 10/10/2022
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: HURTADO

DATE/TIME OF TEST: 10/19/2020 00:13:33

Diagnostics Check		OK	00:10	
Air Blank		0.000	00:10	
Control Test		0.081	00:10	
Air Blank		0.000	00:11	
Subject Sample	# 1		00:12	Subject Test Refused
Breath Volume		0.000	00:12	
Air Blank		0.000	00:12	
Control Test		0.081	00:12	
Air Blank		0.000	00:13	
Diagnostics Check		OK	00:13	

Data Download Date/Time: 11/19/2020 11:25

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HRADECKY,CORBIN,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HRADECKY,CORBIN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 10/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1217042
DRY GAS EXPIRY DATE: 10/10/2022
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: MORELLI III

DATE/TIME OF TEST: 10/21/2020 03:36:39

Diagnostics Check		OK	03:32	
Air Blank		0.000	03:33	
Control Test		0.080	03:33	
Air Blank		0.000	03:34	
Subject Sample	# 1		03:35	Subject Test Refused
Breath Volume		0.000	03:35	
Air Blank		0.000	03:35	
Control Test		0.080	03:36	
Air Blank		0.000	03:36	
Diagnostics Check		OK	03:36	

Data Download Date/Time: 11/19/2020 11:25

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTERO,LISSETTE,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: CRISTEA,VICTOR
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 10/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1217042
DRY GAS EXPIRY DATE: 10/10/2022
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: QUINONES DIAZ

DATE/TIME OF TEST: 10/23/2020 00:56:19

Diagnostics Check		OK	00:53	
Air Blank		0.000	00:53	
Control Test		0.080	00:54	
Air Blank		0.000	00:54	
Subject Sample	# 1		00:54	Subject Test Refused
Breath Volume		0.000	00:54	
Air Blank		0.000	00:55	
Control Test		0.080	00:55	
Air Blank		0.000	00:56	
Diagnostics Check		OK	00:56	

Data Download Date/Time: 11/19/2020 11:25

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTERO,LISSETTE,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MARINO,JUAN CARLOS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 10/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1217042
DRY GAS EXPIRY DATE: 10/10/2022
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: MUNOZ

DATE/TIME OF TEST: 10/23/2020 03:29:36

Diagnostics Check		OK	03:26	
Air Blank		0.000	03:27	
Control Test		0.080	03:27	
Air Blank		0.000	03:28	
Subject Sample	# 1		03:28	Subject Test Refused
Breath Volume		0.000	03:28	
Air Blank		0.000	03:28	
Control Test		0.080	03:28	
Air Blank		0.000	03:29	
Diagnostics Check		OK	03:29	

Data Download Date/Time: 11/19/2020 11:25

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUERRA,WILLIAM,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HRADECKY,CORBIN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 10/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1217042
DRY GAS EXPIRY DATE: 10/10/2022
VIOLATION CODE: DUI
ARREST TIME: 17:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:17

SUBJECT'S LAST NAME: BERGLUND

DATE/TIME OF TEST: 10/23/2020 18:49:27

Diagnostics Check		OK	18:43
Air Blank		0.000	18:43
Control Test		0.079	18:43
Air Blank		0.000	18:44
Subject Sample	# 1	0.152	18:44
Breath Volume		3.214	18:44
Air Blank		0.000	18:45
Air Blank		0.000	18:47
Subject Sample	# 2	0.158	18:47
Breath Volume		3.640	18:47
Air Blank		0.000	18:48
Control Test		0.079	18:48
Air Blank		0.000	18:49
Diagnostics Check		OK	18:49

Data Download Date/Time: 11/19/2020 11:25

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUERRA,WILLIAM,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HRADECKY,CORBIN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 10/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1217042
DRY GAS EXPIRY DATE: 10/10/2022
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: NOEL

DATE/TIME OF TEST: 10/23/2020 23:27:50

Diagnostics Check		OK	23:24	
Air Blank		0.000	23:25	
Control Test		0.080	23:25	
Air Blank		0.000	23:26	
Subject Sample	# 1		23:26	Subject Test Refused
Breath Volume		0.000	23:26	
Air Blank		0.000	23:26	
Control Test		0.081	23:27	
Air Blank		0.000	23:27	
Diagnostics Check		OK	23:27	

Data Download Date/Time: 11/19/2020 11:25

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUERRA,WILLIAM,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HRADECKY,CORBIN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 10/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1217042
DRY GAS EXPIRY DATE: 10/10/2022
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: VERDIN GONZALEZ

DATE/TIME OF TEST: 10/25/2020 03:50:58

Diagnostics Check		OK	03:43
Air Blank		0.000	03:44
Control Test		0.080	03:44
Air Blank		0.000	03:44
Subject Sample	# 1	0.203	03:45
Breath Volume		2.257	03:45
Air Blank		0.000	03:46
Air Blank		0.000	03:48
Subject Sample	# 2	0.199	03:49
Breath Volume		2.328	03:49
Air Blank		0.000	03:50
Control Test		0.079	03:50
Air Blank		0.000	03:50
Diagnostics Check		OK	03:50

Data Download Date/Time: 11/19/2020 11:25

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEIRD,ANDREW,D
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: AGUANNO,CHRISTOPHER
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 10/07/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1217042
DRY GAS EXPIRY DATE: 10/10/2022
VIOLATION CODE: DUI
ARREST TIME: 14:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:33

SUBJECT'S LAST NAME: FELLOWS

DATE/TIME OF TEST: 10/26/2020 16:00:14

Diagnostics Check		OK	15:53
Air Blank		0.000	15:54
Control Test		0.081	15:54
Air Blank		0.000	15:55
Subject Sample	# 1	0.021	15:55
Breath Volume		4.253	15:55
Air Blank		0.000	15:56
Air Blank		0.000	15:58
Subject Sample	# 2	0.018	15:58
Breath Volume		2.054	15:58
Air Blank		0.000	15:59
Control Test		0.080	15:59
Air Blank		0.000	16:00
Diagnostics Check		OK	16:00

Data Download Date/Time: 11/19/2020 11:25

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAFFERTY,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: JEAN-FRANCOIS,LEEF
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 10/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: DOWNEY

DATE/TIME OF TEST: 10/29/2020 01:13:02

Diagnostics Check		OK	00:58	
Air Blank		0.000	00:58	
Control Test		0.080	00:58	
Air Blank		0.000	00:59	
Subject Sample	# 1	0.124	01:02	Volume Not Met
Breath Volume		0.816	01:02	
Air Blank		0.000	01:03	
Air Blank		0.000	01:05	
Subject Sample	# 2	0.139	01:05	
Breath Volume		1.726	01:05	
Air Blank		0.000	01:06	
Air Blank		0.000	01:08	
Subject Sample	# 3	0.120	01:11	Volume Not Met
Breath Volume		0.167	01:11	
Air Blank		0.000	01:12	
Control Test		0.080	01:12	
Air Blank		0.000	01:12	
Diagnostics Check		OK	01:13	

Data Download Date/Time: 11/18/2020 14:52

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GROSSO,DYLAN,J
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: GROSSO,DYLAN
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 10/28/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 19:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:38

SUBJECT'S LAST NAME: ARAGONES

DATE/TIME OF TEST: 10/31/2020 20:04:37

Diagnostics Check		OK	19:58
Air Blank		0.000	19:59
Control Test		0.081	19:59
Air Blank		0.000	20:00
Subject Sample	# 1	0.000	20:00
Breath Volume		2.785	20:00
Air Blank		0.000	20:00
Air Blank		0.000	20:02
Subject Sample	# 2	0.000	20:03
Breath Volume		3.101	20:03
Air Blank		0.000	20:03
Control Test		0.081	20:03
Air Blank		0.000	20:04
Diagnostics Check		OK	20:04

Data Download Date/Time: 11/18/2020 14:52

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLANTON,CRYSTAL,G
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: THEIS,ZACHARY
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 09/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: CIELESZ

DATE/TIME OF TEST: 10/02/2020 00:27:45

Diagnostics Check		OK	00:21
Air Blank		0.000	00:21
Control Test		0.081	00:22
Air Blank		0.000	00:22
Subject Sample	# 1	0.134	00:23
Breath Volume		2.132	00:23
Air Blank		0.000	00:23
Air Blank		0.000	00:25
Subject Sample	# 2	0.134	00:26
Breath Volume		1.613	00:26
Air Blank		0.000	00:26
Control Test		0.079	00:27
Air Blank		0.000	00:27
Diagnostics Check		OK	00:27

Data Download Date/Time: 10/28/2020 12:32

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PULFORD,G,E
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: PULFORD,G
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 09/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 10/02/2020 23:08:34

Diagnostics Check		OK	23:03	
Air Blank		0.000	23:03	
Control Test		0.080	23:04	
Air Blank		0.000	23:04	
Subject Sample	# 1	0.237	23:07	
Breath Volume		1.355	23:07	
Air Blank			23:08	Purge Fail
Air Blank		0.000	23:08	

Data Download Date/Time: 10/28/2020 12:32

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PULFORD,G,E
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: PULFORD,G
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 09/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 10/02/2020 23:18:46

Diagnostics Check		OK	23:11
Air Blank		0.000	23:11
Control Test		0.077	23:11
Air Blank		0.000	23:12
Subject Sample	# 1	0.231	23:13
Breath Volume		2.269	23:13
Air Blank		0.000	23:14
Air Blank		0.000	23:15
Subject Sample	# 2	0.220	23:16
Breath Volume		1.617	23:16
Air Blank		0.000	23:17
Control Test		0.076	23:18
Air Blank		0.000	23:18
Diagnostics Check		OK	23:18

Data Download Date/Time: 10/28/2020 12:32

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,MICAH,
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MOORE,MICAH
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 09/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: BAGGETT

DATE/TIME OF TEST: 10/03/2020 02:09:08

Diagnostics Check		OK	02:02
Air Blank		0.000	02:03
Control Test		0.079	02:03
Air Blank		0.000	02:03
Subject Sample	# 1	0.151	02:04
Breath Volume		3.632	02:04
Air Blank		0.000	02:05
Air Blank		0.000	02:06
Subject Sample	# 2	0.149	02:07
Breath Volume		3.605	02:07
Air Blank		0.000	02:08
Control Test		0.079	02:08
Air Blank		0.000	02:09
Diagnostics Check		OK	02:09

Data Download Date/Time: 10/28/2020 12:32

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSSO,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: RUSSO,CHRISTOPHER
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 09/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 04:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: RUIZ

DATE/TIME OF TEST: 10/03/2020 05:12:14

Diagnostics Check		OK	05:04
Air Blank		0.000	05:04
Control Test		0.080	05:05
Air Blank		0.000	05:05
Subject Sample	# 1	0.182	05:06
Breath Volume		1.679	05:06
Air Blank		0.000	05:07
Air Blank		0.000	05:09
Subject Sample	# 2	0.184	05:10
Breath Volume		1.203	05:10
Air Blank		0.000	05:11
Control Test		0.078	05:11
Air Blank		0.000	05:12
Diagnostics Check		OK	05:12

Data Download Date/Time: 10/28/2020 12:32

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSSO,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: RUSSO,CHRISTOPHER
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 09/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:38

SUBJECT'S LAST NAME: SCHWEBEL

DATE/TIME OF TEST: 10/03/2020 21:08:54

Diagnostics Check		OK	21:01
Air Blank		0.000	21:02
Control Test		0.080	21:02
Air Blank		0.000	21:03
Subject Sample	# 1	0.175	21:03
Breath Volume		1.773	21:03
Air Blank		0.000	21:04
Air Blank		0.000	21:06
Subject Sample	# 2	0.160	21:07
Breath Volume		1.121	21:07
Air Blank		0.000	21:07
Control Test		0.078	21:08
Air Blank		0.000	21:08
Diagnostics Check		OK	21:08

Data Download Date/Time: 10/28/2020 12:32

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOEDICKE,WILLIAM,R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: JOEDICKE,WILLIAM
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 09/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: GAUFFREAU

DATE/TIME OF TEST: 10/04/2020 02:16:42

Diagnostics Check		OK	02:10
Air Blank		0.000	02:10
Control Test		0.079	02:11
Air Blank		0.000	02:11
Subject Sample	# 1	0.252	02:12
Breath Volume		3.695	02:12
Air Blank		0.000	02:12
Air Blank		0.000	02:14
Subject Sample	# 2	0.236	02:14
Breath Volume		3.050	02:14
Air Blank		0.000	02:15
Control Test		0.080	02:16
Air Blank		0.000	02:16
Diagnostics Check		OK	02:16

Data Download Date/Time: 10/28/2020 12:32

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOOREHEAD,DAVID,L
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: MOOREHEAD,DAVID
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 09/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: TELLO CIFUENTES

DATE/TIME OF TEST: 10/04/2020 02:29:08

Diagnostics Check		OK	02:22
Air Blank		0.000	02:22
Control Test		0.080	02:22
Air Blank		0.000	02:23
Subject Sample	# 1	0.182	02:24
Breath Volume		3.078	02:24
Air Blank		0.000	02:25
Air Blank		0.000	02:26
Subject Sample	# 2	0.175	02:27
Breath Volume		2.335	02:27
Air Blank		0.000	02:28
Control Test		0.079	02:28
Air Blank		0.000	02:29
Diagnostics Check		OK	02:29

Data Download Date/Time: 10/28/2020 12:32

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAFFERTY,C,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: JEAN-FRANCOIS,LEEF
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 09/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: FUENTES

DATE/TIME OF TEST: 10/11/2020 02:16:56

Diagnostics Check		OK	02:10
Air Blank		0.000	02:11
Control Test		0.081	02:11
Air Blank		0.000	02:11
Subject Sample	# 1	0.045	02:12
Breath Volume		1.386	02:12
Air Blank		0.000	02:12
Air Blank		0.000	02:14
Subject Sample	# 2	0.045	02:15
Breath Volume		1.531	02:15
Air Blank		0.000	02:15
Control Test		0.082	02:16
Air Blank		0.000	02:16
Diagnostics Check		OK	02:16

Data Download Date/Time: 10/28/2020 12:32

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BALL,CODY,W
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BALL,CODY
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: BORESTKY

DATE/TIME OF TEST: 10/11/2020 03:28:21

Diagnostics Check		OK	03:21
Air Blank		0.000	03:22
Control Test		0.081	03:22
Air Blank		0.000	03:23
Subject Sample	# 1	0.122	03:23
Breath Volume		1.878	03:23
Air Blank		0.000	03:24
Air Blank		0.000	03:26
Subject Sample	# 2	0.117	03:26
Breath Volume		2.015	03:26
Air Blank		0.000	03:27
Control Test		0.080	03:27
Air Blank		0.000	03:28
Diagnostics Check		OK	03:28

Data Download Date/Time: 10/28/2020 12:32

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYS,MATTHEW,W
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: COLLAZO,FERDINAND
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 09/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 10/11/2020 23:45:43

Diagnostics Check		OK	23:38
Air Blank		0.000	23:39
Control Test		0.080	23:39
Air Blank		0.000	23:39
Subject Sample	# 1	0.193	23:40
Breath Volume		2.320	23:40
Air Blank		0.000	23:41
Air Blank		0.000	23:43
Subject Sample	# 2	0.184	23:44
Breath Volume		2.328	23:44
Air Blank		0.000	23:44
Control Test		0.079	23:45
Air Blank		0.000	23:45
Diagnostics Check		OK	23:45

Data Download Date/Time: 10/28/2020 12:32

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DECKER,JORDAN,A
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: MASSEO,DEREK
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 09/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 10:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:20

SUBJECT'S LAST NAME: STRONG

DATE/TIME OF TEST: 10/12/2020 11:11:34

Diagnostics Check		OK	11:02
Air Blank		0.000	11:03
Control Test		0.080	11:03
Air Blank		0.000	11:04
Subject Sample	# 1	0.000	11:07
Breath Volume		2.875	11:07
Air Blank		0.000	11:07
Air Blank		0.000	11:09
Subject Sample	# 2	0.000	11:10
Breath Volume		2.468	11:10
Air Blank		0.000	11:10
Control Test		0.080	11:10
Air Blank		0.000	11:11
Diagnostics Check		OK	11:11

Data Download Date/Time: 10/28/2020 12:32

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERSON,KIP,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: PETERSON,KIP
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 09/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 10/17/2020 00:26:54

Diagnostics Check		OK	00:21	
Air Blank		0.000	00:21	
Control Test		0.081	00:21	
Air Blank		0.000	00:22	
Subject Sample	# 1	0.174	00:25	Volume Not Met
Breath Volume		0.117	00:25	
Air Blank			00:26	Purge Fail
Air Blank		0.000	00:26	

Data Download Date/Time: 10/28/2020 12:32

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERSON,KIP,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: PETERSON,KIP
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 09/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 10/17/2020 00:33:46

Diagnostics Check		OK	00:28	
Air Blank		0.000	00:28	
Control Test		0.078	00:28	
Air Blank		0.000	00:29	
Subject Sample	# 1		00:31	Subject Test Refused
Breath Volume		0.031	00:31	
Air Blank		0.000	00:32	
Control Test		0.079	00:33	
Air Blank		0.000	00:33	
Diagnostics Check		OK	00:33	

Data Download Date/Time: 10/28/2020 12:32

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSACI,NEIL,R
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: ROSACI,NEIL
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 09/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 17:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:37

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 10/17/2020 19:05:59

Diagnostics Check		OK	18:57
Air Blank		0.000	18:58
Control Test		0.080	18:58
Air Blank		0.000	18:59
Subject Sample	# 1	0.247	19:01
Breath Volume		1.535	19:01
Air Blank		0.000	19:02
Air Blank		0.000	19:03
Subject Sample	# 2	0.257	19:04
Breath Volume		1.386	19:04
Air Blank		0.000	19:05
Control Test		0.077	19:05
Air Blank		0.000	19:05
Diagnostics Check		OK	19:05

Data Download Date/Time: 10/28/2020 12:32

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREDETTE,MICHAEL,G
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: FREDETTE,MICHAEL
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 09/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: RAS

DATE/TIME OF TEST: 10/17/2020 22:07:40

Diagnostics Check		OK	22:00
Air Blank		0.000	22:01
Control Test		0.078	22:01
Air Blank		0.000	22:02
Subject Sample	# 1	0.196	22:02
Breath Volume		3.316	22:02
Air Blank		0.000	22:03
Air Blank		0.000	22:05
Subject Sample	# 2	0.196	22:05
Breath Volume		3.406	22:05
Air Blank		0.000	22:06
Control Test		0.079	22:07
Air Blank		0.000	22:07
Diagnostics Check		OK	22:07

Data Download Date/Time: 10/28/2020 12:32

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTOS,JASON,L
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: O'GRADY,SHELBY
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 09/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 08:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:20

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 10/19/2020 08:32:27

Diagnostics Check	Fail	08:31	Internal Printer Test
Air Blank	0.000	08:32	

Data Download Date/Time: 10/28/2020 12:32

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTOS,JASON,L
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: O'GRADY,SHELBY
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 09/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 08:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:20

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 10/19/2020 08:35:24

Diagnostics Check	Fail	08:34	Internal Printer Test
Air Blank	0.000	08:35	

Data Download Date/Time: 10/28/2020 12:32

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTOS,JASON,L
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: O'GRADY,SHELBY
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 09/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 08:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:20

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 10/19/2020 09:00:09

Diagnostics Check		OK	08:53
Air Blank		0.000	08:54
Control Test		0.080	08:54
Air Blank		0.000	08:54
Subject Sample	# 1	0.387	08:55
Breath Volume		2.988	08:55
Air Blank		0.000	08:56
Air Blank		0.000	08:57
Subject Sample	# 2	0.371	08:58
Breath Volume		2.875	08:58
Air Blank		0.000	08:59
Control Test		0.078	08:59
Air Blank		0.000	09:00
Diagnostics Check		OK	09:00

Data Download Date/Time: 10/28/2020 12:32

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BALL,CODY,W
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BALL,CODY
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 19:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 10/21/2020 20:21:54

Diagnostics Check		OK	20:07	
Air Blank		0.000	20:08	
Control Test		0.081	20:08	
Air Blank		0.000	20:09	
Subject Sample	# 1		20:12	No Sample Provided
Breath Volume		0.000	20:12	
Air Blank		0.000	20:12	
Air Blank		0.000	20:14	
Subject Sample	# 2	0.157	20:16	
Breath Volume		1.789	20:16	
Air Blank		0.000	20:17	
Air Blank		0.000	20:19	
Subject Sample	# 3	0.165	20:20	
Breath Volume		1.789	20:20	
Air Blank		0.000	20:20	
Control Test		0.080	20:21	
Air Blank		0.000	20:21	
Diagnostics Check		OK	20:21	

Data Download Date/Time: 10/28/2020 12:32

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MERCER,ISAAC,V
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: MERCER,ISAAC
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 09/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: OUELLETTE

DATE/TIME OF TEST: 10/22/2020 02:58:50

Diagnostics Check		OK	02:52
Air Blank		0.000	02:52
Control Test		0.081	02:52
Air Blank		0.000	02:53
Subject Sample	# 1	0.177	02:53
Breath Volume		3.695	02:53
Air Blank		0.000	02:54
Air Blank		0.000	02:56
Subject Sample	# 2	0.178	02:57
Breath Volume		2.312	02:57
Air Blank		0.000	02:57
Control Test		0.079	02:58
Air Blank		0.000	02:58
Diagnostics Check		OK	02:58

Data Download Date/Time: 10/28/2020 12:32

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 09/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: CACERES

DATE/TIME OF TEST: 10/23/2020 22:42:41

Diagnostics Check		OK	22:32
Air Blank		0.000	22:33
Control Test		0.081	22:33
Air Blank		0.000	22:33
Subject Sample	# 1	0.112	22:36
Breath Volume		2.566	22:36
Air Blank		0.000	22:37
Air Blank		0.000	22:39
Subject Sample	# 2	0.115	22:40
Breath Volume		3.054	22:40
Air Blank		0.000	22:41
Control Test		0.080	22:42
Air Blank		0.000	22:42
Diagnostics Check		OK	22:42

Data Download Date/Time: 10/28/2020 12:32

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWMAN,BRYAN,D
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: BOWMAN,BRYAN
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 09/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 12:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: UNDERWOOD

DATE/TIME OF TEST: 10/24/2020 01:41:42

Diagnostics Check		OK	01:35
Air Blank		0.000	01:35
Control Test		0.080	01:36
Air Blank		0.000	01:36
Subject Sample	# 1	0.037	01:37
Breath Volume		2.234	01:37
Air Blank		0.000	01:37
Air Blank		0.000	01:39
Subject Sample	# 2	0.034	01:40
Breath Volume		1.574	01:40
Air Blank		0.000	01:40
Control Test		0.080	01:41
Air Blank		0.000	01:41
Diagnostics Check		OK	01:41

Data Download Date/Time: 10/28/2020 12:32

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYS,MATTHEW,W
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: PAZ,CHRISTOPHER
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 09/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: FLOWERS

DATE/TIME OF TEST: 10/25/2020 01:44:06

Diagnostics Check		OK	01:37
Air Blank		0.000	01:38
Control Test		0.081	01:38
Air Blank		0.000	01:38
Subject Sample	# 1	0.175	01:39
Breath Volume		1.507	01:39
Air Blank		0.000	01:40
Air Blank		0.000	01:41
Subject Sample	# 2	0.174	01:42
Breath Volume		1.843	01:42
Air Blank		0.000	01:43
Control Test		0.081	01:43
Air Blank		0.000	01:44
Diagnostics Check		OK	01:44

Data Download Date/Time: 10/28/2020 12:32

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHULTZ,JEREMY,J
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: DUNAWAY,SAMUEL
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 09/18/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: HENDERSON

DATE/TIME OF TEST: 10/07/2020 02:28:09

Diagnostics Check		OK	02:16	
Air Blank		0.000	02:17	
Control Test		0.079	02:17	
Air Blank		0.000	02:17	
Subject Sample	# 1		02:21	No Sample Provided
Breath Volume		0.000	02:21	
Air Blank		0.000	02:21	
Air Blank		0.000	02:23	
Subject Sample	# 2		02:26	No Sample Provided
Breath Volume		0.000	02:26	
Air Blank		0.000	02:27	
Control Test		0.081	02:27	
Air Blank		0.000	02:28	
Diagnostics Check		OK	02:28	

Data Download Date/Time: 10/16/2020 23:25

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COKONOUGHER,ROBERT,F
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: COKONOUGHER,ROBERT
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 10/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: DRY

DATE/TIME OF TEST: 10/21/2020 02:26:28

Diagnostics Check	Fail	02:26	Internal Printer Test
Air Blank	0.000	02:26	

Data Download Date/Time: 11/24/2020 12:08

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COKONOUGHER,ROBERT,F
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: COKONOUGHER,ROBERT
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 10/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: DRY

DATE/TIME OF TEST: 10/21/2020 02:39:30

Diagnostics Check		OK	02:27	
Air Blank		0.000	02:28	
Control Test		0.080	02:28	
Air Blank		0.000	02:29	
Subject Sample	# 1		02:32	No Sample Provided
Breath Volume		0.000	02:32	
Air Blank		0.000	02:32	
Air Blank		0.000	02:34	
Subject Sample	# 2		02:37	No Sample Provided
Breath Volume		0.000	02:37	
Air Blank		0.000	02:38	
Control Test		0.080	02:38	
Air Blank		0.000	02:39	
Diagnostics Check		OK	02:39	

Data Download Date/Time: 11/24/2020 12:08

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURGESS,MASON,W
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BURGESS,MASON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/16/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 14:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:30

SUBJECT'S LAST NAME: BURGESS

DATE/TIME OF TEST: 10/22/2020 16:42:58

Diagnostics Check		OK	16:31	
Air Blank		0.000	16:32	
Control Test		0.078	16:32	
Air Blank		0.000	16:33	
Subject Sample	# 1		16:36	No Sample Provided
Breath Volume		0.000	16:36	
Air Blank		0.000	16:36	
Air Blank		0.000	16:38	
Subject Sample	# 2		16:40	Subject Test Refused
Breath Volume		0.000	16:40	
Air Blank		0.000	16:41	
Control Test		0.080	16:41	
Air Blank		0.000	16:42	
Diagnostics Check		Fail	16:42	Internal Printer Test
Air Blank		0.000	16:42	

Data Download Date/Time: 11/24/2020 12:12

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AIMANOVICH,MATTHEW,
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: AIMANOVICH,MATTHEW
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 10/16/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: CRUM

DATE/TIME OF TEST: 10/22/2020 23:31:53

Diagnostics Check		OK	23:23	
Air Blank		0.000	23:23	
Control Test		0.078	23:23	
Air Blank		0.000	23:24	
Subject Sample	# 1		23:27	No Sample Provided
Breath Volume		0.000	23:27	
Air Blank		0.000	23:28	
Air Blank		0.000	23:30	
Subject Sample	# 2		23:30	Subject Test Refused
Breath Volume		0.000	23:30	
Air Blank		0.000	23:30	
Control Test		0.079	23:31	
Air Blank		0.000	23:31	
Diagnostics Check		OK	23:31	

Data Download Date/Time: 11/24/2020 12:12

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSE,ALEXIS,M
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: MAY,WILLIAM
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 10/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: WISE

DATE/TIME OF TEST: 10/25/2020 02:25:13

Diagnostics Check		OK	02:18
Air Blank		0.000	02:19
Control Test		0.077	02:19
Air Blank		0.000	02:19
Subject Sample	# 1	0.061	02:20
Breath Volume		2.695	02:20
Air Blank		0.000	02:21
Air Blank		0.000	02:23
Subject Sample	# 2	0.062	02:23
Breath Volume		3.175	02:23
Air Blank		0.000	02:24
Control Test		0.081	02:24
Air Blank		0.000	02:25
Diagnostics Check		OK	02:25

Data Download Date/Time: 11/24/2020 12:12

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARQUIS,NORMAND,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MARQUIS,NORMAND
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: HATAWAY

DATE/TIME OF TEST: 10/28/2020 02:07:18

Diagnostics Check		OK	01:58	
Air Blank		0.000	01:59	
Control Test		0.079	01:59	
Air Blank		0.000	02:00	
Subject Sample	# 1	0.194	02:02	
Breath Volume		1.390	02:02	
Air Blank		0.000	02:03	
Air Blank		0.000	02:04	
Subject Sample	# 2	0.177	02:05	
Breath Volume		2.027	02:05	
Air Blank		0.000	02:06	
Control Test		0.060	02:06	Control Outside Tolerance
Air Blank		0.000	02:07	

Data Download Date/Time: 11/24/2020 12:12

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARQUIS,NORMAND,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MARQUIS,NORMAND
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 12:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: THAMES

DATE/TIME OF TEST: 10/29/2020 02:11:44

Diagnostics Check		OK	02:02
Air Blank		0.000	02:02
Control Test		0.076	02:02
Air Blank		0.000	02:03
Subject Sample	# 1	0.185	02:06
Breath Volume		2.769	02:06
Air Blank		0.000	02:06
Air Blank		0.000	02:08
Subject Sample	# 2	0.172	02:09
Breath Volume		2.746	02:09
Air Blank		0.000	02:10
Control Test		0.080	02:11
Air Blank		0.000	02:11
Diagnostics Check		OK	02:11

Data Download Date/Time: 11/24/2020 12:12

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRASER,MICHAEL,C
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: BYRD,BRANDON
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1160200
DRY GAS EXPIRY DATE: 06/20/2022
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: ORR

DATE/TIME OF TEST: 10/02/2020 03:07:14

Diagnostics Check		OK	03:00
Air Blank		0.000	03:00
Control Test		0.079	03:00
Air Blank		0.000	03:01
Subject Sample	# 1	0.200	03:02
Breath Volume		1.714	03:02
Air Blank		0.000	03:03
Air Blank		0.000	03:04
Subject Sample	# 2	0.198	03:05
Breath Volume		1.648	03:05
Air Blank		0.000	03:06
Control Test		0.076	03:06
Air Blank		0.000	03:07
Diagnostics Check		OK	03:07

Data Download Date/Time: 11/09/2020 14:44

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,BRENDAN,D
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: BROWN,BRENDAN
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1160200
DRY GAS EXPIRY DATE: 06/20/2022
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: LORRAINE

DATE/TIME OF TEST: 10/03/2020 02:05:06

Diagnostics Check		OK	01:58
Air Blank		0.000	01:58
Control Test		0.079	01:58
Air Blank		0.000	01:59
Subject Sample	# 1	0.199	02:00
Breath Volume		3.101	02:00
Air Blank		0.000	02:01
Air Blank		0.000	02:02
Subject Sample	# 2	0.202	02:03
Breath Volume		4.187	02:03
Air Blank		0.000	02:04
Control Test		0.077	02:04
Air Blank		0.000	02:05
Diagnostics Check		OK	02:05

Data Download Date/Time: 11/09/2020 14:44

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUDSON,JOSHUA,K
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: HUDSON,JOSHUA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1160200
DRY GAS EXPIRY DATE: 06/20/2022
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 10/03/2020 04:04:58

Diagnostics Check		OK	03:58
Air Blank		0.000	03:58
Control Test		0.079	03:59
Air Blank		0.000	03:59
Subject Sample	# 1	0.161	04:00
Breath Volume		2.539	04:00
Air Blank		0.000	04:01
Air Blank		0.000	04:02
Subject Sample	# 2	0.163	04:03
Breath Volume		2.433	04:03
Air Blank		0.000	04:04
Control Test		0.078	04:04
Air Blank		0.000	04:04
Diagnostics Check		OK	04:04

Data Download Date/Time: 11/09/2020 14:44

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,BRENDAN,D
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: ROPER,WILLIAM
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1160200
DRY GAS EXPIRY DATE: 06/20/2022
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: SHRAUGER

DATE/TIME OF TEST: 10/04/2020 02:31:26

Diagnostics Check		OK	02:29	
Air Blank		0.000	02:29	
Control Test		0.079	02:30	
Air Blank		0.000	02:30	
Subject Sample	# 1		02:30	Sequence Aborted
Breath Volume		0.000	02:30	
Air Blank		0.000	02:31	

Data Download Date/Time: 11/13/2020 08:26

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,BRENDAN,D
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: ROPER,WILLIAM
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1160200
DRY GAS EXPIRY DATE: 06/20/2022
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: SHRAUGER

DATE/TIME OF TEST: 10/04/2020 03:00:20

Diagnostics Check		OK	02:53
Air Blank		0.000	02:53
Control Test		0.079	02:54
Air Blank		0.000	02:54
Subject Sample	# 1	0.185	02:55
Breath Volume		2.406	02:55
Air Blank		0.000	02:56
Air Blank		0.000	02:57
Subject Sample	# 2	0.180	02:58
Breath Volume		3.000	02:58
Air Blank		0.000	02:59
Control Test		0.078	02:59
Air Blank		0.000	03:00
Diagnostics Check		OK	03:00

Data Download Date/Time: 11/13/2020 08:26

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMAHON,KENNETH,J
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: MLINAC,MADISON
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1160200
DRY GAS EXPIRY DATE: 06/20/2022
VIOLATION CODE: DUI
ARREST TIME: 04:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: SELLERS

DATE/TIME OF TEST: 10/06/2020 05:19:59

Diagnostics Check		OK	05:13
Air Blank		0.000	05:13
Control Test		0.079	05:14
Air Blank		0.000	05:14
Subject Sample	# 1	0.000	05:15
Breath Volume		2.207	05:15
Air Blank		0.000	05:16
Air Blank		0.000	05:18
Subject Sample	# 2	0.000	05:18
Breath Volume		1.882	05:18
Air Blank		0.000	05:19
Control Test		0.079	05:19
Air Blank		0.000	05:19
Diagnostics Check		OK	05:19

Data Download Date/Time: 11/13/2020 08:26

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMAHON,KENNETH,J
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: ARGENT,BRANDON
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1160200
DRY GAS EXPIRY DATE: 06/20/2022
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: DINIZ

DATE/TIME OF TEST: 10/10/2020 03:37:33

Diagnostics Check		OK	03:31
Air Blank		0.000	03:31
Control Test		0.078	03:31
Air Blank		0.000	03:32
Subject Sample	# 1	0.141	03:32
Breath Volume		1.371	03:32
Air Blank		0.000	03:33
Air Blank		0.000	03:35
Subject Sample	# 2	0.139	03:35
Breath Volume		2.496	03:35
Air Blank		0.000	03:36
Control Test		0.079	03:36
Air Blank		0.000	03:37
Diagnostics Check		OK	03:37

Data Download Date/Time: 11/13/2020 08:26

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMAHON,KENNETH,J
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: COTTON,PIERCE
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1160200
DRY GAS EXPIRY DATE: 06/20/2022
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: MCWILLIAMS

DATE/TIME OF TEST: 10/10/2020 04:21:10

Diagnostics Check		OK	04:15
Air Blank		0.000	04:15
Control Test		0.079	04:16
Air Blank		0.000	04:16
Subject Sample	# 1	0.000	04:16
Breath Volume		1.855	04:16
Air Blank		0.000	04:17
Air Blank		0.000	04:19
Subject Sample	# 2	0.000	04:19
Breath Volume		2.484	04:19
Air Blank		0.000	04:20
Control Test		0.079	04:20
Air Blank		0.000	04:21
Diagnostics Check		OK	04:21

Data Download Date/Time: 11/13/2020 08:26

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TASCHETTA,JOSEPH,S
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: MLINAC,MADISON
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1160200
DRY GAS EXPIRY DATE: 06/20/2022
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: STRICKLAND

DATE/TIME OF TEST: 10/11/2020 03:10:24

Diagnostics Check		OK	03:03
Air Blank		0.000	03:04
Control Test		0.079	03:04
Air Blank		0.000	03:05
Subject Sample	# 1	0.221	03:05
Breath Volume		2.210	03:05
Air Blank		0.000	03:06
Air Blank		0.000	03:08
Subject Sample	# 2	0.215	03:08
Breath Volume		2.097	03:08
Air Blank		0.000	03:09
Control Test		0.078	03:09
Air Blank		0.000	03:10
Diagnostics Check		OK	03:10

Data Download Date/Time: 11/13/2020 08:26

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,JAMIE,L
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: NELSON,JAMIE
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1160200
DRY GAS EXPIRY DATE: 06/20/2022
VIOLATION CODE: DUI
ARREST TIME: 21:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:16

SUBJECT'S LAST NAME: WELCH

DATE/TIME OF TEST: 10/12/2020 21:50:00

Diagnostics Check		OK	21:43
Air Blank		0.000	21:44
Control Test		0.080	21:44
Air Blank		0.000	21:45
Subject Sample	# 1	0.213	21:45
Breath Volume		1.460	21:45
Air Blank		0.000	21:46
Air Blank		0.000	21:47
Subject Sample	# 2	0.228	21:48
Breath Volume		2.863	21:48
Air Blank		0.000	21:49
Control Test		0.078	21:49
Air Blank		0.000	21:49
Diagnostics Check		OK	21:50

Data Download Date/Time: 11/13/2020 08:26

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEACH,ZACHARY,M
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: BYRD,BRANDON
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1160200
DRY GAS EXPIRY DATE: 06/20/2022
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 10/16/2020 04:23:15

Diagnostics Check		OK	04:14
Air Blank		0.000	04:14
Control Test		0.079	04:14
Air Blank		0.000	04:15
Subject Sample	# 1	0.158	04:17
Breath Volume		1.703	04:17
Air Blank		0.000	04:18
Air Blank		0.000	04:20
Subject Sample	# 2	0.148	04:21
Breath Volume		2.257	04:21
Air Blank		0.000	04:22
Control Test		0.075	04:22
Air Blank		0.000	04:23
Diagnostics Check		OK	04:23

Data Download Date/Time: 11/13/2020 08:26

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,BRENDAN,D
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: BROWN,BRENDAN
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/28/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1160200
DRY GAS EXPIRY DATE: 06/20/2022
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: HASENBERG

DATE/TIME OF TEST: 10/17/2020 01:48:46

Diagnostics Check		OK	01:41
Air Blank		0.000	01:42
Control Test		0.080	01:42
Air Blank		0.000	01:43
Subject Sample	# 1	0.000	01:44
Breath Volume		4.058	01:44
Air Blank		0.000	01:44
Air Blank		0.000	01:46
Subject Sample	# 2	0.000	01:47
Breath Volume		3.179	01:47
Air Blank		0.000	01:47
Control Test		0.080	01:48
Air Blank		0.000	01:48
Diagnostics Check		OK	01:48

Data Download Date/Time: 11/13/2020 08:26

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,JAMIE,L
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: NELSON,JAMIE
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1160200
DRY GAS EXPIRY DATE: 06/20/2022
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 10/18/2020 04:36:51

Diagnostics Check		OK	04:31
Air Blank		0.000	04:31
Control Test		0.079	04:31
Air Blank		0.000	04:32
Subject Sample	# 1	0.000	04:32
Breath Volume		2.234	04:32
Air Blank		0.000	04:33
Air Blank		0.000	04:35
Subject Sample	# 2	0.000	04:35
Breath Volume		2.492	04:35
Air Blank		0.000	04:35
Control Test		0.079	04:36
Air Blank		0.000	04:36
Diagnostics Check		OK	04:36

Data Download Date/Time: 11/13/2020 08:26

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLLINS,CHASE,M
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: COLLINS,CHASE
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1160200
DRY GAS EXPIRY DATE: 06/20/2022
VIOLATION CODE: DUI
ARREST TIME: 05:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:12

SUBJECT'S LAST NAME: FALCON

DATE/TIME OF TEST: 10/18/2020 06:39:25

Diagnostics Check		OK	06:33
Air Blank		0.000	06:33
Control Test		0.080	06:34
Air Blank		0.000	06:34
Subject Sample	# 1	0.153	06:34
Breath Volume		2.296	06:34
Air Blank		0.000	06:35
Air Blank		0.000	06:37
Subject Sample	# 2	0.152	06:37
Breath Volume		2.515	06:37
Air Blank		0.000	06:38
Control Test		0.079	06:38
Air Blank		0.000	06:39
Diagnostics Check		OK	06:39

Data Download Date/Time: 11/13/2020 08:26

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUDSON,JOSHUA,K
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: PINO,RICHARD
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1160200
DRY GAS EXPIRY DATE: 06/20/2022
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: DEMARIA

DATE/TIME OF TEST: 10/21/2020 00:01:56

Diagnostics Check		OK	23:53
Air Blank		0.000	23:54
Control Test		0.079	23:54
Air Blank		0.000	23:55
Subject Sample	# 2	0.175	23:56
Breath Volume		1.718	23:56
Air Blank		0.000	23:57
Air Blank		0.000	23:59
Subject Sample	# 1	0.163	00:00
Breath Volume		1.761	00:00
Air Blank		0.000	00:01
Control Test		0.078	00:01
Air Blank		0.000	00:01
Diagnostics Check		OK	00:01

Data Download Date/Time: 11/13/2020 08:26

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINDBLOM,ROBERT,A
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: LINDBLOM,ROBERT
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1160200
DRY GAS EXPIRY DATE: 06/20/2022
VIOLATION CODE: DUI
ARREST TIME: 05:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:52

SUBJECT'S LAST NAME: MCCOY

DATE/TIME OF TEST: 10/22/2020 06:20:44

Diagnostics Check		OK	06:14
Air Blank		0.000	06:14
Control Test		0.079	06:15
Air Blank		0.000	06:15
Subject Sample	# 1	0.201	06:16
Breath Volume		1.753	06:16
Air Blank		0.000	06:16
Air Blank		0.000	06:18
Subject Sample	# 2	0.196	06:19
Breath Volume		3.355	06:19
Air Blank		0.000	06:19
Control Test		0.078	06:20
Air Blank		0.000	06:20
Diagnostics Check		OK	06:20

Data Download Date/Time: 11/13/2020 08:26

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEACH,ZACHARY,M
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: LEACH,ZACHARY
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1160200
DRY GAS EXPIRY DATE: 06/20/2022
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: ENCISO

DATE/TIME OF TEST: 10/24/2020 00:19:17

Diagnostics Check		OK	00:06	
Air Blank		0.000	00:06	
Control Test		0.079	00:07	
Air Blank		0.000	00:07	
Subject Sample	# 1	0.233	00:08	No .020 agreement
Breath Volume		2.394	00:08	
Air Blank		0.000	00:09	
Air Blank		0.000	00:10	
Subject Sample	# 2	0.204	00:12	No .020 agreement
Breath Volume		1.957	00:12	
Air Blank		0.000	00:12	
Air Blank		0.000	00:14	
Subject Sample	# 3	0.191	00:17	Volume Not Met
Breath Volume		0.054	00:17	
Air Blank		0.000	00:18	
Control Test		0.075	00:18	
Air Blank		0.000	00:19	
Diagnostics Check		OK	00:19	

Data Download Date/Time: 11/13/2020 08:26

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUDSON,JOSHUA,K
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: PINO,RICHARD
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1160200
DRY GAS EXPIRY DATE: 06/20/2022
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:37

SUBJECT'S LAST NAME: BOSARREYES

DATE/TIME OF TEST: 10/24/2020 22:05:37

Diagnostics Check		OK	21:59
Air Blank		0.000	21:59
Control Test		0.080	22:00
Air Blank		0.000	22:00
Subject Sample	# 1	0.121	22:01
Breath Volume		3.488	22:01
Air Blank		0.000	22:01
Air Blank		0.000	22:03
Subject Sample	# 2	0.123	22:04
Breath Volume		3.207	22:04
Air Blank		0.000	22:04
Control Test		0.079	22:05
Air Blank		0.000	22:05
Diagnostics Check		OK	22:05

Data Download Date/Time: 11/13/2020 08:26

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TASCHETTA,JOSEPH,S
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: GLASSMAN,GARRETT
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1160200
DRY GAS EXPIRY DATE: 06/20/2022
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: RIETZ

DATE/TIME OF TEST: 10/24/2020 23:02:33

Diagnostics Check		OK	22:59	
Air Blank		0.000	22:59	
Control Test		0.079	23:00	
Air Blank		0.000	23:00	
Subject Sample	# 1		23:00	Slope Not Met
Breath Volume		1.523	23:00	
Air Blank		0.000	23:01	
Control Test		0.078	23:01	
Air Blank		0.000	23:02	
Diagnostics Check		OK	23:02	

Data Download Date/Time: 11/13/2020 08:26

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TASCHETTA,JOSEPH,S
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: GLASSMAN,GARRETT
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1160200
DRY GAS EXPIRY DATE: 06/20/2022
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: RIETZ

DATE/TIME OF TEST: 10/24/2020 23:10:35

Diagnostics Check		OK	23:04
Air Blank		0.000	23:04
Control Test		0.078	23:05
Air Blank		0.000	23:05
Subject Sample	# 1	0.253	23:06
Breath Volume		2.773	23:06
Air Blank		0.000	23:06
Air Blank		0.000	23:08
Subject Sample	# 2	0.243	23:09
Breath Volume		1.566	23:09
Air Blank		0.000	23:09
Control Test		0.077	23:09
Air Blank		0.000	23:10
Diagnostics Check		OK	23:10

Data Download Date/Time: 11/13/2020 08:26

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BYRD,BRANDON,C
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: BYRD,BRANDON
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 10/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1160200
DRY GAS EXPIRY DATE: 06/20/2022
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: SPEED

DATE/TIME OF TEST: 10/30/2020 02:45:33

Diagnostics Check	OK	02:44	
Air Blank	0.000	02:44	
Control Test	0.079	02:44	
Air Blank		02:45	RFI Detect
Air Blank	0.000	02:45	

Data Download Date/Time: 11/13/2020 08:26

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BYRD,BRANDON,C
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: BYRD,BRANDON
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 10/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1160200
DRY GAS EXPIRY DATE: 06/20/2022
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: SPEED

DATE/TIME OF TEST: 10/30/2020 02:54:32

Diagnostics Check		OK	02:47
Air Blank		0.000	02:47
Control Test		0.079	02:48
Air Blank		0.000	02:48
Subject Sample	# 1	0.150	02:49
Breath Volume		2.093	02:49
Air Blank		0.000	02:50
Air Blank		0.000	02:51
Subject Sample	# 2	0.152	02:52
Breath Volume		3.265	02:52
Air Blank		0.000	02:53
Control Test		0.078	02:53
Air Blank		0.000	02:54
Diagnostics Check		OK	02:54

Data Download Date/Time: 11/13/2020 08:26

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORREA,JONATHAN,E
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: CORREA,JONATHAN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 09/21/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: HEBENSTREIT

DATE/TIME OF TEST: 10/02/2020 01:22:25

Diagnostics Check		OK	01:15
Air Blank		0.000	01:16
Control Test		0.079	01:16
Air Blank		0.000	01:16
Subject Sample	# 1	0.177	01:17
Breath Volume		1.433	01:17
Air Blank		0.000	01:18
Air Blank		0.000	01:19
Subject Sample	# 2	0.182	01:20
Breath Volume		1.820	01:20
Air Blank		0.000	01:21
Control Test		0.078	01:21
Air Blank		0.000	01:22
Diagnostics Check		OK	01:22

Data Download Date/Time: 10/19/2020 11:47

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUSCIO,JESSE,J
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: CRAWFORD,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: HARRIGAN

DATE/TIME OF TEST: 10/04/2020 03:28:13

Diagnostics Check		OK	03:22
Air Blank		0.000	03:22
Control Test		0.079	03:22
Air Blank		0.000	03:23
Subject Sample	# 1	0.137	03:23
Breath Volume		1.761	03:23
Air Blank		0.000	03:24
Air Blank		0.000	03:26
Subject Sample	# 2	0.126	03:26
Breath Volume		1.449	03:26
Air Blank		0.000	03:27
Control Test		0.077	03:27
Air Blank		0.000	03:28
Diagnostics Check		OK	03:28

Data Download Date/Time: 10/19/2020 11:47

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUSCIO,JESSE,J
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: TRUSCIO,JESSE
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 19:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: UHLAND

DATE/TIME OF TEST: 10/04/2020 20:53:57

Diagnostics Check		OK	20:50	
Air Blank		0.000	20:50	
Control Test		0.079	20:50	
Air Blank		0.000	20:51	
Subject Sample	# 1		20:52	Subject Test Refused
Breath Volume		0.000	20:52	
Air Blank		0.000	20:53	
Control Test		0.079	20:53	
Air Blank		0.000	20:53	
Diagnostics Check		OK	20:53	

Data Download Date/Time: 10/19/2020 11:47

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONIE,JARED,P
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: MONIE,JARED
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:12

SUBJECT'S LAST NAME: HANSON

DATE/TIME OF TEST: 10/04/2020 21:36:49

Diagnostics Check		OK	21:33	
Air Blank		0.000	21:34	
Control Test		0.079	21:34	
Air Blank		0.000	21:35	
Subject Sample	# 1		21:35	Subject Test Refused
Breath Volume		0.000	21:35	
Air Blank		0.000	21:35	
Control Test		0.078	21:36	
Air Blank		0.000	21:36	
Diagnostics Check		OK	21:36	

Data Download Date/Time: 10/19/2020 11:47

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUSCIO,JESSE,J
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: TRUSCIO,JESSE
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: HALES

DATE/TIME OF TEST: 10/08/2020 23:27:17

Diagnostics Check		OK	23:20
Air Blank		0.000	23:21
Control Test		0.078	23:21
Air Blank		0.000	23:22
Subject Sample	# 1	0.244	23:22
Breath Volume		1.464	23:22
Air Blank		0.000	23:23
Air Blank		0.000	23:25
Subject Sample	# 2	0.231	23:25
Breath Volume		1.351	23:25
Air Blank		0.000	23:26
Control Test		0.078	23:26
Air Blank		0.000	23:27
Diagnostics Check		OK	23:27

Data Download Date/Time: 10/19/2020 11:47

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: SYED

DATE/TIME OF TEST: 10/15/2020 02:07:00

Diagnostics Check		OK	02:00
Air Blank		0.000	02:01
Control Test		0.079	02:01
Air Blank		0.000	02:02
Subject Sample	# 1	0.157	02:02
Breath Volume		1.964	02:02
Air Blank		0.000	02:03
Air Blank		0.000	02:05
Subject Sample	# 2	0.150	02:05
Breath Volume		1.539	02:05
Air Blank		0.000	02:06
Control Test		0.078	02:06
Air Blank		0.000	02:06
Diagnostics Check		OK	02:07

Data Download Date/Time: 10/19/2020 11:47

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: STUART,S
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: WYSONG

DATE/TIME OF TEST: 10/15/2020 03:53:21

Diagnostics Check		OK	03:47
Air Blank		0.000	03:47
Control Test		0.078	03:47
Air Blank		0.000	03:48
Subject Sample	# 1	0.236	03:48
Breath Volume		1.699	03:48
Air Blank		0.000	03:49
Air Blank		0.000	03:51
Subject Sample	# 2	0.229	03:51
Breath Volume		1.289	03:51
Air Blank		0.000	03:52
Control Test		0.078	03:52
Air Blank		0.000	03:53
Diagnostics Check		OK	03:53

Data Download Date/Time: 10/19/2020 11:47

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: ROGERSON

DATE/TIME OF TEST: 10/15/2020 22:43:05

Diagnostics Check	OK	22:42	
Air Blank		22:42	RFI Detect
Air Blank	0.000	22:43	

Data Download Date/Time: 10/19/2020 11:47

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: ROGERSON

DATE/TIME OF TEST: 10/15/2020 22:47:05

Diagnostics Check		OK	22:44	
Air Blank		0.000	22:44	
Control Test		0.079	22:45	
Air Blank		0.000	22:45	
Subject Sample	# 1		22:45	Subject Test Refused
Breath Volume		0.000	22:45	
Air Blank		0.000	22:46	
Control Test		0.078	22:46	
Air Blank		0.000	22:47	
Diagnostics Check		OK	22:47	

Data Download Date/Time: 10/19/2020 11:47

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,DARBY,J
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: FLOYD,JOHN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: DRURY

DATE/TIME OF TEST: 10/17/2020 01:24:22

Diagnostics Check		OK	01:21	
Air Blank		0.000	01:21	
Control Test		0.078	01:22	
Air Blank		0.000	01:22	
Subject Sample	# 1		01:23	Subject Test Refused
Breath Volume		0.000	01:23	
Air Blank		0.000	01:23	
Control Test		0.079	01:23	
Air Blank		0.000	01:24	
Diagnostics Check		OK	01:24	

Data Download Date/Time: 10/19/2020 11:47

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,DARBY,J
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: CORDLE,ABIGAIL
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 09/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 10/17/2020 01:45:45

Diagnostics Check		OK	01:39
Air Blank		0.000	01:40
Control Test		0.079	01:40
Air Blank		0.000	01:40
Subject Sample	# 1	0.153	01:41
Breath Volume		2.683	01:41
Air Blank		0.000	01:41
Air Blank		0.000	01:43
Subject Sample	# 2	0.152	01:44
Breath Volume		2.617	01:44
Air Blank		0.000	01:44
Control Test		0.078	01:45
Air Blank		0.000	01:45
Diagnostics Check		OK	01:45

Data Download Date/Time: 10/19/2020 11:48

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GELSEY,RICHARD,D
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: BROWN,BENJAMIN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: GRIFFITHS

DATE/TIME OF TEST: 10/18/2020 03:38:14

Diagnostics Check		OK	03:31
Air Blank		0.000	03:31
Control Test		0.079	03:32
Air Blank		0.000	03:32
Subject Sample	# 1	0.161	03:33
Breath Volume		1.804	03:33
Air Blank		0.000	03:34
Air Blank		0.000	03:36
Subject Sample	# 2	0.165	03:36
Breath Volume		1.988	03:36
Air Blank		0.000	03:37
Control Test		0.079	03:37
Air Blank		0.000	03:38
Diagnostics Check		OK	03:38

Data Download Date/Time: 10/19/2020 11:48

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 10/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: CALHOUN

DATE/TIME OF TEST: 10/20/2020 03:04:58

Diagnostics Check		OK	02:58
Air Blank		0.000	02:59
Control Test		0.077	02:59
Air Blank		0.000	03:00
Subject Sample	# 1	0.095	03:00
Breath Volume		2.316	03:00
Air Blank		0.000	03:01
Air Blank		0.000	03:03
Subject Sample	# 2	0.094	03:03
Breath Volume		2.019	03:03
Air Blank		0.000	03:04
Control Test		0.077	03:04
Air Blank		0.000	03:04
Diagnostics Check		OK	03:04

Data Download Date/Time: 11/12/2020 10:01

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,DARBY,J
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: BROWN,BENJAMIN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 10/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: ADDISON

DATE/TIME OF TEST: 10/22/2020 01:17:02

Diagnostics Check		OK	01:14	
Air Blank		0.000	01:14	
Control Test		0.077	01:15	
Air Blank		0.000	01:15	
Subject Sample	# 1		01:15	Subject Test Refused
Breath Volume		0.000	01:15	
Air Blank		0.000	01:16	
Control Test		0.077	01:16	
Air Blank		0.000	01:16	
Diagnostics Check		OK	01:17	

Data Download Date/Time: 11/12/2020 10:01

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,DARBY,J
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: CORDLE,ABIGAIL
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 10/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: LEARY

DATE/TIME OF TEST: 10/23/2020 03:03:33

Diagnostics Check		OK	03:00	
Air Blank		0.000	03:01	
Control Test		0.077	03:01	
Air Blank		0.000	03:02	
Subject Sample	# 1		03:02	Subject Test Refused
Breath Volume		0.000	03:02	
Air Blank		0.000	03:02	
Control Test		0.077	03:02	
Air Blank		0.000	03:03	
Diagnostics Check		OK	03:03	

Data Download Date/Time: 11/12/2020 10:01

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAYWARD,KYLE,M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: PEARCE,BRANDON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 10/19/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 12:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: VENEGAS

DATE/TIME OF TEST: 10/24/2020 01:49:02

Diagnostics Check		OK	01:42
Air Blank		0.000	01:43
Control Test		0.077	01:43
Air Blank		0.000	01:44
Subject Sample	# 1	0.168	01:44
Breath Volume		1.195	01:44
Air Blank		0.000	01:45
Air Blank		0.000	01:47
Subject Sample	# 2	0.161	01:47
Breath Volume		1.109	01:47
Air Blank		0.000	01:48
Control Test		0.076	01:48
Air Blank		0.000	01:48
Diagnostics Check		OK	01:49

Data Download Date/Time: 11/12/2020 10:01

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 10/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: EDENFIELD

DATE/TIME OF TEST: 10/24/2020 04:00:07

Diagnostics Check		OK	03:57	
Air Blank		0.000	03:57	
Control Test		0.077	03:58	
Air Blank		0.000	03:58	
Subject Sample	# 1		03:58	Subject Test Refused
Breath Volume		0.000	03:58	
Air Blank		0.000	03:59	
Control Test		0.077	03:59	
Air Blank		0.000	04:00	
Diagnostics Check		OK	04:00	

Data Download Date/Time: 11/12/2020 10:01

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAYWARD,KYLE,M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: POWELL,AARON
ARRESTING OFFICER AGENCY: St. Augustine Beach
LAST AGENCY INSPECTION: 10/19/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: WILDER

DATE/TIME OF TEST: 10/25/2020 02:21:13

Diagnostics Check		OK	02:17	
Air Blank		0.000	02:18	
Control Test		0.078	02:18	
Air Blank		0.000	02:19	
Subject Sample	# 1		02:19	Subject Test Refused
Breath Volume		0.000	02:19	
Air Blank		0.000	02:20	
Control Test		0.078	02:20	
Air Blank		0.000	02:21	
Diagnostics Check		OK	02:21	

Data Download Date/Time: 11/12/2020 10:01

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAYWARD,KYLE,M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: POWELL,AARON
ARRESTING OFFICER AGENCY: St. Augustine Beach
LAST AGENCY INSPECTION: 10/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: BRICKER

DATE/TIME OF TEST: 10/26/2020 00:23:13

Diagnostics Check		OK	00:17	
Air Blank		0.000	00:17	
Control Test		0.077	00:17	
Air Blank		0.000	00:18	
Subject Sample	# 1	0.227	00:18	
Breath Volume		2.531	00:18	
Air Blank		0.000	00:19	
Air Blank		0.000	00:21	
Subject Sample	# 2	0.239	00:21	
Breath Volume		2.765	00:21	
Air Blank		0.000	00:22	
Control Test		0.074	00:22	Control Outside Tolerance
Air Blank		0.000	00:23	

Data Download Date/Time: 11/12/2020 10:01

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YARBOROUGH,ERIC,L
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: BARRERA,ALEXANDER
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 10/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: DOVER

DATE/TIME OF TEST: 10/27/2020 02:49:02

Diagnostics Check		OK	02:42
Air Blank		0.000	02:43
Control Test		0.078	02:43
Air Blank		0.000	02:43
Subject Sample	# 1	0.131	02:44
Breath Volume		2.285	02:44
Air Blank		0.000	02:45
Air Blank		0.000	02:47
Subject Sample	# 2	0.138	02:47
Breath Volume		2.390	02:47
Air Blank		0.000	02:48
Control Test		0.077	02:48
Air Blank		0.000	02:48
Diagnostics Check		OK	02:49

Data Download Date/Time: 11/12/2020 10:01

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: JAROUSSE,D
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 10/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 11:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:17

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 10/29/2020 12:54:51

Diagnostics Check		OK	12:44
Air Blank		0.000	12:44
Control Test		0.078	12:44
Air Blank		0.000	12:45
Subject Sample	# 1	0.274	12:46
Breath Volume		1.832	12:46
Air Blank		0.000	12:46
Air Blank		0.000	12:48
Subject Sample	# 2	0.253	12:50
Breath Volume		2.039	12:50
Air Blank		0.000	12:50
Air Blank		0.000	12:52
Subject Sample	# 3	0.255	12:53
Breath Volume		2.175	12:53
Air Blank		0.000	12:53
Control Test		0.076	12:54
Air Blank		0.000	12:54
Diagnostics Check		OK	12:54

Data Download Date/Time: 11/12/2020 10:01

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AYERS,SPENCER,O
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: CUBBEDGE,KYLE
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 10/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: TALTON

DATE/TIME OF TEST: 10/29/2020 22:19:40

Diagnostics Check		OK	22:11	
Air Blank		0.000	22:11	
Control Test		0.077	22:11	
Air Blank		0.000	22:12	
Subject Sample	# 1		22:15	No Sample Provided
Breath Volume		0.000	22:15	
Air Blank		0.000	22:16	
Air Blank		0.000	22:18	
Subject Sample	# 2		22:18	Subject Test Refused
Breath Volume		0.000	22:18	
Air Blank		0.000	22:18	
Control Test		0.077	22:19	
Air Blank		0.000	22:19	
Diagnostics Check		OK	22:19	

Data Download Date/Time: 11/12/2020 10:01

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 10/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: SURRENCY

DATE/TIME OF TEST: 10/30/2020 03:17:55

Diagnostics Check		OK	03:11
Air Blank		0.000	03:12
Control Test		0.077	03:12
Air Blank		0.000	03:12
Subject Sample	# 1	0.204	03:13
Breath Volume		2.417	03:13
Air Blank		0.000	03:14
Air Blank		0.000	03:15
Subject Sample	# 2	0.202	03:16
Breath Volume		2.445	03:16
Air Blank		0.000	03:17
Control Test		0.076	03:17
Air Blank		0.000	03:17
Diagnostics Check		OK	03:17

Data Download Date/Time: 11/12/2020 10:01

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUSCIO,JESSE,J
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: TRUSCIO,JESSE
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 10/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:23

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 10/30/2020 05:50:34

Diagnostics Check		OK	05:43
Air Blank		0.000	05:44
Control Test		0.077	05:44
Air Blank		0.000	05:45
Subject Sample	# 1	0.272	05:45
Breath Volume		1.742	05:45
Air Blank		0.000	05:46
Air Blank		0.000	05:48
Subject Sample	# 2	0.269	05:48
Breath Volume		1.984	05:48
Air Blank		0.000	05:49
Control Test		0.076	05:49
Air Blank		0.000	05:50
Diagnostics Check		OK	05:50

Data Download Date/Time: 11/12/2020 10:01

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMATEER,TRACY,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: PALMATEER,TRACY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/18/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: TAGESSON

DATE/TIME OF TEST: 10/04/2020 03:28:45

Diagnostics Check		OK	03:22
Air Blank		0.000	03:22
Control Test		0.077	03:23
Air Blank		0.000	03:23
Subject Sample	# 1	0.121	03:24
Breath Volume		1.300	03:24
Air Blank		0.000	03:24
Air Blank		0.000	03:26
Subject Sample	# 2	0.127	03:27
Breath Volume		2.175	03:27
Air Blank		0.000	03:27
Control Test		0.075	03:28
Air Blank		0.000	03:28
Diagnostics Check		OK	03:28

Data Download Date/Time: 11/24/2020 10:16

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FUGATE,MITCHELL,W
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: CREAMER,REMINGTON
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 09/18/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:38

SUBJECT'S LAST NAME: ROSEN

DATE/TIME OF TEST: 10/04/2020 19:03:35

Diagnostics Check		OK	18:59	
Air Blank		0.000	18:59	
Control Test		0.077	18:59	
Air Blank		0.000	19:00	
Subject Sample	# 1		19:02	Subject Test Refused
Breath Volume		0.000	19:02	
Air Blank		0.000	19:02	
Control Test		0.077	19:02	
Air Blank		0.000	19:03	
Diagnostics Check		OK	19:03	

Data Download Date/Time: 11/24/2020 10:16

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENNETT,DAVID,J
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: CREAMER,REMINGTON
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 09/18/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: WAGUESPACK

DATE/TIME OF TEST: 10/05/2020 18:53:30

Diagnostics Check		OK	18:47
Air Blank		0.000	18:47
Control Test		0.077	18:48
Air Blank		0.000	18:48
Subject Sample	# 1	0.160	18:49
Breath Volume		3.367	18:49
Air Blank		0.000	18:49
Air Blank		0.000	18:51
Subject Sample	# 2	0.157	18:51
Breath Volume		3.285	18:51
Air Blank		0.000	18:52
Control Test		0.076	18:52
Air Blank		0.000	18:53
Diagnostics Check		OK	18:53

Data Download Date/Time: 11/24/2020 10:16

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENNETT,DAVID,J
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: RAYA,ZACHARY
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 09/18/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: LONG

DATE/TIME OF TEST: 10/10/2020 23:21:03

Diagnostics Check		OK	23:16	
Air Blank		0.000	23:16	
Control Test		0.077	23:17	
Air Blank		0.000	23:17	
Subject Sample	# 1		23:19	Subject Test Refused
Breath Volume		0.000	23:19	
Air Blank		0.000	23:20	
Control Test		0.077	23:20	
Air Blank		0.000	23:20	
Diagnostics Check		OK	23:21	

Data Download Date/Time: 11/24/2020 10:16

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSE,ALEXIS,M
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: SCALIA,ADAM
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 09/18/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 10/15/2020 01:01:42

Diagnostics Check		OK	00:54
Air Blank		0.000	00:54
Control Test		0.076	00:55
Air Blank		0.000	00:55
Subject Sample	# 1	0.165	00:56
Breath Volume		2.777	00:56
Air Blank		0.000	00:57
Air Blank		0.000	00:59
Subject Sample	# 2	0.156	01:00
Breath Volume		2.730	01:00
Air Blank		0.000	01:00
Control Test		0.076	01:01
Air Blank		0.000	01:01
Diagnostics Check		OK	01:01

Data Download Date/Time: 11/24/2020 10:16

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BOONE,M
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 09/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: DELIVRA

DATE/TIME OF TEST: 10/05/2020 00:53:12

Diagnostics Check		OK	00:46
Air Blank		0.000	00:46
Control Test		0.082	00:47
Air Blank		0.000	00:47
Subject Sample	# 1	0.153	00:48
Breath Volume		2.191	00:48
Air Blank		0.000	00:49
Air Blank		0.000	00:51
Subject Sample	# 2	0.155	00:51
Breath Volume		2.152	00:51
Air Blank		0.000	00:52
Control Test		0.080	00:52
Air Blank		0.000	00:53
Diagnostics Check		OK	00:53

Data Download Date/Time: 10/08/2020 11:04

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: THOMAS,JOSEPH
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 09/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: FUENTES SUAREZ

DATE/TIME OF TEST: 10/05/2020 04:53:28

Diagnostics Check		OK	04:45
Air Blank		0.000	04:45
Control Test		0.081	04:46
Air Blank		0.000	04:46
Subject Sample	# 1	0.160	04:47
Breath Volume		1.214	04:47
Air Blank		0.000	04:48
Air Blank		0.000	04:49
Subject Sample	# 2	0.159	04:51
Breath Volume		1.292	04:51
Air Blank		0.000	04:52
Control Test		0.081	04:52
Air Blank		0.000	04:53
Diagnostics Check		OK	04:53

Data Download Date/Time: 10/08/2020 11:04

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: STUEHRENBERG,ERIK
ARRESTING OFFICER AGENCY: Davie PD
LAST AGENCY INSPECTION: 09/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: BADILLO

DATE/TIME OF TEST: 10/06/2020 04:07:38

Diagnostics Check		OK	04:01
Air Blank		0.000	04:01
Control Test		0.081	04:01
Air Blank		0.000	04:02
Subject Sample	# 1	0.122	04:02
Breath Volume		2.281	04:02
Air Blank		0.000	04:03
Air Blank		0.000	04:05
Subject Sample	# 2	0.123	04:06
Breath Volume		1.238	04:06
Air Blank		0.000	04:06
Control Test		0.081	04:07
Air Blank		0.000	04:07
Diagnostics Check		OK	04:07

Data Download Date/Time: 10/08/2020 11:04

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HOLLOWAY,DEVARIOUS
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 09/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 06:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:03

SUBJECT'S LAST NAME: SIDDONS

DATE/TIME OF TEST: 10/07/2020 08:39:11

Diagnostics Check		OK	08:27	
Air Blank		0.000	08:28	
Control Test		0.082	08:28	
Air Blank		0.000	08:29	
Subject Sample	# 1		08:32	No Sample Provided
Breath Volume		0.000	08:32	
Air Blank		0.000	08:33	
Air Blank		0.000	08:34	
Subject Sample	# 2		08:37	No Sample Provided
Breath Volume		0.000	08:37	
Air Blank		0.000	08:38	
Control Test		0.081	08:38	
Air Blank		0.000	08:39	
Diagnostics Check		OK	08:39	

Data Download Date/Time: 10/08/2020 11:04

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARMSTRONG,BRITTANY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ARMSTRONG,BRITTANY
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 09/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 20:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:07

SUBJECT'S LAST NAME: HOUSLEY

DATE/TIME OF TEST: 10/07/2020 21:42:58

Diagnostics Check		OK	21:35
Air Blank		0.000	21:36
Control Test		0.081	21:36
Air Blank		0.000	21:37
Subject Sample	# 1	0.272	21:37
Breath Volume		1.574	21:37
Air Blank		0.000	21:38
Air Blank		0.000	21:40
Subject Sample	# 2	0.272	21:41
Breath Volume		1.878	21:41
Air Blank		0.000	21:42
Control Test		0.081	21:42
Air Blank		0.000	21:42
Diagnostics Check		OK	21:42

Data Download Date/Time: 10/08/2020 11:04

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HOLLIGAN,DANIEL
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 09/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 10/08/2020 05:39:21

Diagnostics Check		OK	05:30
Air Blank		0.000	05:30
Control Test		0.081	05:31
Air Blank		0.000	05:31
Subject Sample	# 1	0.216	05:34
Breath Volume		1.316	05:34
Air Blank		0.000	05:35
Air Blank		0.000	05:37
Subject Sample	# 2	0.202	05:37
Breath Volume		1.945	05:37
Air Blank		0.000	05:38
Control Test		0.081	05:38
Air Blank		0.000	05:39
Diagnostics Check		OK	05:39

Data Download Date/Time: 10/08/2020 11:04

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RAUL,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: GILPIN,SCHELISA
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 09/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: GODINEZ LOPEZ

DATE/TIME OF TEST: 10/03/2020 03:51:22

Diagnostics Check		OK	03:42	
Air Blank		0.000	03:43	
Control Test		0.083	03:43	
Air Blank		0.000	03:44	
Subject Sample	# 1		03:47	No Sample Provided
Breath Volume		0.000	03:47	
Air Blank		0.000	03:47	
Air Blank		0.000	03:49	
Subject Sample	# 2		03:49	Subject Test Refused
Breath Volume		0.000	03:49	
Air Blank		0.000	03:50	
Control Test		0.082	03:50	
Air Blank		0.000	03:51	
Diagnostics Check		OK	03:51	

Data Download Date/Time: 11/02/2020 07:58

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,WAYNE,L
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: ROGERS,WAYNE
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 09/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 10/03/2020 22:35:25

Diagnostics Check		OK	22:27
Air Blank		0.000	22:27
Control Test		0.082	22:28
Air Blank		0.000	22:28
Subject Sample	# 1	0.276	22:30
Breath Volume		1.757	22:30
Air Blank		0.000	22:31
Air Blank		0.000	22:32
Subject Sample	# 2	0.273	22:33
Breath Volume		2.257	22:33
Air Blank		0.000	22:34
Control Test		0.080	22:34
Air Blank		0.000	22:35
Diagnostics Check		OK	22:35

Data Download Date/Time: 11/02/2020 07:58

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILPIN,SCHELISA,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: GILPIN,SCHELISA
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 09/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: LOPEZ SALES

DATE/TIME OF TEST: 10/08/2020 02:14:13

Diagnostics Check		OK	02:04	
Air Blank		0.000	02:05	
Control Test		0.082	02:05	
Air Blank		0.000	02:06	
Subject Sample	# 1		02:09	No Sample Provided
Breath Volume		0.000	02:09	
Air Blank		0.000	02:09	
Air Blank		0.000	02:11	
Subject Sample	# 2		02:12	Subject Test Refused
Breath Volume		0.000	02:12	
Air Blank		0.000	02:13	
Control Test		0.081	02:13	
Air Blank		0.000	02:14	
Diagnostics Check		OK	02:14	

Data Download Date/Time: 11/02/2020 07:58

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: PEREZ,RAUL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 09/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: DOMINGO PEREZ

DATE/TIME OF TEST: 10/09/2020 22:03:50

Diagnostics Check		OK	21:56
Air Blank		0.000	21:57
Control Test		0.082	21:57
Air Blank		0.000	21:58
Subject Sample	# 1	0.257	21:58
Breath Volume		1.671	21:58
Air Blank		0.000	21:59
Air Blank		0.000	22:01
Subject Sample	# 2	0.259	22:01
Breath Volume		2.250	22:01
Air Blank		0.000	22:02
Control Test		0.079	22:03
Air Blank		0.000	22:03
Diagnostics Check		OK	22:03

Data Download Date/Time: 11/02/2020 07:58

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILPIN,SCHELISA,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: GILPIN,SCHELISA
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 09/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: MILIAN AREVALO

DATE/TIME OF TEST: 10/10/2020 04:01:37

Diagnostics Check		OK	03:53	
Air Blank		0.000	03:53	
Control Test		0.081	03:53	
Air Blank		0.000	03:54	
Subject Sample	# 1	0.124	03:57	Volume Not Met
Breath Volume		0.238	03:57	
Air Blank		0.000	03:58	
Air Blank		0.000	03:59	
Subject Sample	# 2		04:00	Subject Test Refused
Breath Volume		0.000	04:00	
Air Blank		0.000	04:00	
Control Test		0.077	04:00	
Air Blank		0.000	04:01	
Diagnostics Check		OK	04:01	

Data Download Date/Time: 11/02/2020 07:58

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: PEREZ,RAUL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 09/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 21:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:13

SUBJECT'S LAST NAME: SOLORZANO SANTOS

DATE/TIME OF TEST: 10/16/2020 21:46:30

Diagnostics Check		OK	21:39
Air Blank		0.000	21:40
Control Test		0.082	21:40
Air Blank		0.000	21:41
Subject Sample	# 1	0.084	21:41
Breath Volume		2.214	21:41
Air Blank		0.000	21:42
Air Blank		0.000	21:44
Subject Sample	# 2	0.076	21:44
Breath Volume		1.539	21:44
Air Blank		0.000	21:45
Control Test		0.081	21:45
Air Blank		0.000	21:46
Diagnostics Check		OK	21:46

Data Download Date/Time: 11/02/2020 07:58

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: PEREZ,RAUL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 09/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: GONZALEZ MUNOS

DATE/TIME OF TEST: 10/17/2020 01:01:52

Diagnostics Check		OK	00:54
Air Blank		0.000	00:55
Control Test		0.081	00:55
Air Blank		0.000	00:56
Subject Sample	# 1	0.185	00:56
Breath Volume		2.734	00:56
Air Blank		0.000	00:57
Air Blank		0.000	00:59
Subject Sample	# 2	0.183	01:00
Breath Volume		2.199	01:00
Air Blank		0.000	01:00
Control Test		0.081	01:01
Air Blank		0.000	01:01
Diagnostics Check		OK	01:01

Data Download Date/Time: 11/02/2020 07:58

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILPIN,SCHELISA,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: GILPIN,SCHELISA
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 09/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: ERREGUIN JIMENEZ

DATE/TIME OF TEST: 10/19/2020 02:52:58

Diagnostics Check		OK	02:45
Air Blank		0.000	02:46
Control Test		0.082	02:46
Air Blank		0.000	02:47
Subject Sample	# 1	0.000	02:48
Breath Volume		1.921	02:48
Air Blank		0.000	02:48
Air Blank		0.000	02:50
Subject Sample	# 2	0.000	02:51
Breath Volume		2.664	02:51
Air Blank		0.000	02:52
Control Test		0.082	02:52
Air Blank		0.000	02:52
Diagnostics Check		OK	02:52

Data Download Date/Time: 11/02/2020 07:58

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: PEREZ,RAUL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 09/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: SANCHEZ GONZALEZ

DATE/TIME OF TEST: 10/21/2020 22:24:49

Diagnostics Check		OK	22:16	
Air Blank		0.000	22:16	
Control Test		0.082	22:17	
Air Blank		0.000	22:17	
Subject Sample	# 1		22:20	No Sample Provided
Breath Volume		0.000	22:20	
Air Blank		0.000	22:21	
Air Blank		0.000	22:23	
Subject Sample	# 2		22:23	Subject Test Refused
Breath Volume		0.000	22:23	
Air Blank		0.000	22:23	
Control Test		0.081	22:24	
Air Blank		0.000	22:24	
Diagnostics Check		OK	22:24	

Data Download Date/Time: 11/02/2020 07:58

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIANCHI,MARIO,D
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: BIANCHI,MARIO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: SABAJMENDOZA

DATE/TIME OF TEST: 10/28/2020 01:38:42

Diagnostics Check		OK	01:31
Air Blank		0.000	01:31
Control Test		0.082	01:32
Air Blank		0.000	01:32
Subject Sample	# 1	0.170	01:33
Breath Volume		1.281	01:33
Air Blank		0.000	01:34
Air Blank		0.000	01:35
Subject Sample	# 2	0.163	01:36
Breath Volume		1.117	01:36
Air Blank		0.000	01:37
Control Test		0.081	01:38
Air Blank		0.000	01:38
Diagnostics Check		OK	01:38

Data Download Date/Time: 11/02/2020 07:58

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILPIN,SCHELISA,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: MARTINEZ,SYLVIA
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 10/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1237851
DRY GAS EXPIRY DATE: 11/21/2022
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: GODINEZ JIMENEZ

DATE/TIME OF TEST: 10/31/2020 04:39:24

Diagnostics Check		OK	04:30	
Air Blank		0.000	04:31	
Control Test		0.081	04:31	
Air Blank		0.000	04:31	
Subject Sample	# 1		04:35	No Sample Provided
Breath Volume		0.000	04:35	
Air Blank		0.000	04:35	
Air Blank		0.000	04:37	
Subject Sample	# 2		04:37	Subject Test Refused
Breath Volume		0.000	04:37	
Air Blank		0.000	04:38	
Control Test		0.080	04:38	
Air Blank		0.000	04:39	
Diagnostics Check		OK	04:39	

Data Download Date/Time: 11/02/2020 07:58

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MORGAN,DAVID
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 10/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: BUTTERS

DATE/TIME OF TEST: 10/10/2020 01:05:13

Diagnostics Check		OK	00:57
Air Blank		0.000	00:57
Control Test		0.080	00:58
Air Blank		0.000	00:58
Subject Sample	# 1	0.230	01:00
Breath Volume		5.226	01:00
Air Blank		0.000	01:01
Air Blank		0.000	01:03
Subject Sample	# 2	0.214	01:03
Breath Volume		1.554	01:03
Air Blank		0.000	01:04
Control Test		0.076	01:04
Air Blank		0.000	01:05
Diagnostics Check		OK	01:05

Data Download Date/Time: 11/23/2020 14:48

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BOHM,JOHN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 10/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: TUNNAGE

DATE/TIME OF TEST: 10/10/2020 04:38:15

Diagnostics Check		OK	04:28
Air Blank		0.000	04:28
Control Test		0.079	04:29
Air Blank		0.000	04:29
Subject Sample	# 1	0.209	04:31
Breath Volume		1.238	04:31
Air Blank		0.000	04:32
Air Blank		0.000	04:34
Subject Sample	# 2	0.214	04:36
Breath Volume		1.371	04:36
Air Blank		0.000	04:37
Control Test		0.076	04:37
Air Blank		0.000	04:38
Diagnostics Check		OK	04:38

Data Download Date/Time: 11/23/2020 14:48

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAPP,JOSHUA,R
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HEPBURN,A
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 10/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 16:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:51

SUBJECT'S LAST NAME: ARCHER

DATE/TIME OF TEST: 10/11/2020 18:22:03

Diagnostics Check		OK	18:14
Air Blank		0.000	18:14
Control Test		0.081	18:15
Air Blank		0.000	18:15
Subject Sample	# 1	0.000	18:17
Breath Volume		1.546	18:17
Air Blank		0.000	18:17
Air Blank		0.000	18:19
Subject Sample	# 2	0.000	18:20
Breath Volume		1.476	18:20
Air Blank		0.000	18:21
Control Test		0.080	18:21
Air Blank		0.000	18:21
Diagnostics Check		OK	18:22

Data Download Date/Time: 11/23/2020 14:48

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: O'DOR,ROBERT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: BUYSENS

DATE/TIME OF TEST: 10/11/2020 23:52:39

Diagnostics Check		OK	23:45
Air Blank		0.000	23:46
Control Test		0.080	23:46
Air Blank		0.000	23:47
Subject Sample	# 1	0.192	23:48
Breath Volume		2.480	23:48
Air Blank		0.000	23:48
Air Blank		0.000	23:50
Subject Sample	# 2	0.174	23:51
Breath Volume		1.199	23:51
Air Blank		0.000	23:51
Control Test		0.079	23:52
Air Blank		0.000	23:52
Diagnostics Check		OK	23:52

Data Download Date/Time: 11/23/2020 14:48

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: TOUSSAINT,LESLIE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 06:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:00

SUBJECT'S LAST NAME: DE PAOLA

DATE/TIME OF TEST: 10/14/2020 08:28:30

Diagnostics Check		OK	08:24	
Air Blank		0.000	08:25	
Control Test		0.079	08:25	
Air Blank		0.000	08:25	
Subject Sample	# 1		08:27	Subject Test Refused
Breath Volume		0.000	08:27	
Air Blank		0.000	08:27	
Control Test		0.081	08:27	
Air Blank		0.000	08:28	
Diagnostics Check		OK	08:28	

Data Download Date/Time: 11/23/2020 14:48

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: TOUSSAINT,LESLIE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 06:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:00

SUBJECT'S LAST NAME: DE PAOLA

DATE/TIME OF TEST: 10/14/2020 08:43:33

Diagnostics Check		OK	08:34
Air Blank		0.000	08:35
Control Test		0.080	08:35
Air Blank		0.000	08:35
Subject Sample	# 1	0.000	08:37
Breath Volume		2.562	08:37
Air Blank		0.000	08:38
Air Blank		0.000	08:40
Subject Sample	# 2	0.000	08:42
Breath Volume		2.503	08:42
Air Blank		0.000	08:42
Control Test		0.079	08:42
Air Blank		0.000	08:43
Diagnostics Check		OK	08:43

Data Download Date/Time: 11/23/2020 14:48

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:47

SUBJECT'S LAST NAME: ZAROS

DATE/TIME OF TEST: 10/16/2020 22:23:55

Diagnostics Check		OK	22:17
Air Blank		0.000	22:17
Control Test		0.080	22:18
Air Blank		0.000	22:18
Subject Sample	# 1	0.190	22:19
Breath Volume		1.699	22:19
Air Blank		0.000	22:19
Air Blank		0.000	22:21
Subject Sample	# 2	0.192	22:22
Breath Volume		1.824	22:22
Air Blank		0.000	22:22
Control Test		0.080	22:23
Air Blank		0.000	22:23
Diagnostics Check		OK	22:23

Data Download Date/Time: 11/23/2020 14:48

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: RODDRIGUEZ,JAMES
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 10/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:47

SUBJECT'S LAST NAME: WILSEY

DATE/TIME OF TEST: 10/17/2020 05:18:40

Diagnostics Check		OK	05:11
Air Blank		0.000	05:11
Control Test		0.080	05:11
Air Blank		0.000	05:12
Subject Sample	# 1	0.201	05:13
Breath Volume		1.707	05:13
Air Blank		0.000	05:14
Air Blank		0.000	05:16
Subject Sample	# 2	0.201	05:16
Breath Volume		1.648	05:16
Air Blank		0.000	05:17
Control Test		0.078	05:18
Air Blank		0.000	05:18
Diagnostics Check		OK	05:18

Data Download Date/Time: 11/23/2020 14:48

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LALANE,LYNN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 10/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: FIGUEROA

DATE/TIME OF TEST: 10/22/2020 23:54:57

Diagnostics Check		OK	23:46
Air Blank		0.000	23:46
Control Test		0.080	23:47
Air Blank		0.000	23:47
Subject Sample	# 1	0.168	23:49
Breath Volume		1.824	23:49
Air Blank		0.000	23:50
Air Blank		0.000	23:52
Subject Sample	# 2	0.179	23:53
Breath Volume		2.878	23:53
Air Blank		0.000	23:54
Control Test		0.077	23:54
Air Blank		0.000	23:54
Diagnostics Check		OK	23:54

Data Download Date/Time: 11/23/2020 14:48

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LUNA,KRISTINA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: LA ROSA

DATE/TIME OF TEST: 10/23/2020 01:56:13

Diagnostics Check		OK	01:49
Air Blank		0.000	01:49
Control Test		0.080	01:49
Air Blank		0.000	01:50
Subject Sample	# 1	0.095	01:51
Breath Volume		1.339	01:51
Air Blank		0.000	01:51
Air Blank		0.000	01:53
Subject Sample	# 2	0.097	01:54
Breath Volume		1.898	01:54
Air Blank		0.000	01:55
Control Test		0.081	01:55
Air Blank		0.000	01:56
Diagnostics Check		OK	01:56

Data Download Date/Time: 11/23/2020 14:48

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: VELEZ,JONATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 10/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 13:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:12

SUBJECT'S LAST NAME: MORGENSTERN

DATE/TIME OF TEST: 10/23/2020 15:44:49

Diagnostics Check		OK	15:38
Air Blank		0.000	15:38
Control Test		0.080	15:39
Air Blank		0.000	15:39
Subject Sample	# 1	0.000	15:40
Breath Volume		2.980	15:40
Air Blank		0.000	15:40
Air Blank		0.000	15:42
Subject Sample	# 2	0.000	15:43
Breath Volume		3.113	15:43
Air Blank		0.000	15:43
Control Test		0.080	15:44
Air Blank		0.000	15:44
Diagnostics Check		OK	15:44

Data Download Date/Time: 11/23/2020 14:48

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Miramar PD
ARRESTING OFFICER: PRESUME,JEHUD
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 10/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: PALACIOS

DATE/TIME OF TEST: 10/24/2020 22:07:44

Diagnostics Check		OK	22:00
Air Blank		0.000	22:01
Control Test		0.081	22:01
Air Blank		0.000	22:02
Subject Sample	# 1	0.000	22:03
Breath Volume		2.207	22:03
Air Blank		0.000	22:03
Air Blank		0.000	22:05
Subject Sample	# 2	0.000	22:06
Breath Volume		2.460	22:06
Air Blank		0.000	22:06
Control Test		0.080	22:07
Air Blank		0.000	22:07
Diagnostics Check		OK	22:07

Data Download Date/Time: 11/23/2020 14:48

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Hollywood PD
ARRESTING OFFICER: MCGUIRE,THOMAS
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 10/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: MENDEZ

DATE/TIME OF TEST: 10/24/2020 22:45:58

Diagnostics Check		OK	22:38
Air Blank		0.000	22:39
Control Test		0.081	22:39
Air Blank		0.000	22:39
Subject Sample	# 1	0.018	22:40
Breath Volume		2.304	22:40
Air Blank		0.000	22:41
Air Blank		0.000	22:43
Subject Sample	# 2	0.017	22:44
Breath Volume		1.218	22:44
Air Blank		0.000	22:45
Control Test		0.081	22:45
Air Blank		0.000	22:45
Diagnostics Check		OK	22:45

Data Download Date/Time: 11/23/2020 14:49

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HERNANDEZ,ERNESTO
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 10/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: VILLABON

DATE/TIME OF TEST: 10/24/2020 23:46:28

Diagnostics Check		OK	23:39
Air Blank		0.000	23:40
Control Test		0.081	23:40
Air Blank		0.000	23:41
Subject Sample	# 1	0.117	23:41
Breath Volume		3.835	23:41
Air Blank		0.000	23:42
Air Blank		0.000	23:44
Subject Sample	# 2	0.111	23:44
Breath Volume		3.925	23:44
Air Blank		0.000	23:45
Control Test		0.080	23:45
Air Blank		0.000	23:46
Diagnostics Check		OK	23:46

Data Download Date/Time: 11/23/2020 14:49

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 04:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:34

SUBJECT'S LAST NAME: ZAMUDIO-SOLORZANO

DATE/TIME OF TEST: 10/25/2020 06:10:55

Diagnostics Check		OK	05:58
Air Blank		0.000	05:59
Control Test		0.081	05:59
Air Blank		0.000	05:59
Subject Sample	# 1	0.186	06:01
Breath Volume		1.929	06:01
Air Blank		0.000	06:02
Air Blank		0.000	06:03
Subject Sample	# 2	0.164	06:05
Breath Volume		1.238	06:05
Air Blank		0.000	06:06
Air Blank		0.000	06:07
Subject Sample	# 3	0.164	06:09
Breath Volume		2.019	06:09
Air Blank		0.000	06:09
Control Test		0.079	06:10
Air Blank		0.000	06:10
Diagnostics Check		OK	06:10

Data Download Date/Time: 11/23/2020 14:49

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HEPBURN,ANDREW
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 10/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 18:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:19

SUBJECT'S LAST NAME: SCRIVEN

DATE/TIME OF TEST: 10/26/2020 19:48:46

Diagnostics Check		OK	19:41
Air Blank		0.000	19:42
Control Test		0.081	19:42
Air Blank		0.000	19:43
Subject Sample	# 1	0.147	19:43
Breath Volume		1.769	19:43
Air Blank		0.000	19:44
Air Blank		0.000	19:46
Subject Sample	# 2	0.146	19:47
Breath Volume		1.796	19:47
Air Blank		0.000	19:47
Control Test		0.080	19:48
Air Blank		0.000	19:48
Diagnostics Check		OK	19:48

Data Download Date/Time: 11/23/2020 14:49

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 10/29/2020 01:59:26

Diagnostics Check		OK	01:47	
Air Blank		0.000	01:47	
Control Test		0.080	01:48	
Air Blank		0.000	01:48	
Subject Sample	# 1		01:51	No Sample Provided
Breath Volume		0.000	01:51	
Air Blank		0.000	01:52	
Air Blank		0.000	01:54	
Subject Sample	# 2	0.241	01:57	Volume Not Met
Breath Volume		0.371	01:57	
Air Blank		0.000	01:58	
Control Test		0.080	01:58	
Air Blank		0.000	01:59	
Diagnostics Check		OK	01:59	

Data Download Date/Time: 11/23/2020 14:49

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 10/29/2020 02:09:48

Diagnostics Check		OK	02:02
Air Blank		0.000	02:02
Control Test		0.078	02:03
Air Blank		0.000	02:03
Subject Sample	# 1	0.262	02:04
Breath Volume		1.312	02:04
Air Blank		0.000	02:05
Air Blank		0.000	02:06
Subject Sample	# 2	0.248	02:07
Breath Volume		1.554	02:07
Air Blank		0.000	02:08
Control Test		0.080	02:09
Air Blank		0.000	02:09
Diagnostics Check		OK	02:09

Data Download Date/Time: 11/23/2020 14:49

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: SHEKHTER

DATE/TIME OF TEST: 10/29/2020 04:27:42

Diagnostics Check		OK	04:21
Air Blank		0.000	04:21
Control Test		0.080	04:22
Air Blank		0.000	04:22
Subject Sample	# 1	0.000	04:23
Breath Volume		1.183	04:23
Air Blank		0.000	04:23
Air Blank		0.000	04:25
Subject Sample	# 2	0.000	04:26
Breath Volume		3.074	04:26
Air Blank		0.000	04:26
Control Test		0.080	04:27
Air Blank		0.000	04:27
Diagnostics Check		OK	04:27

Data Download Date/Time: 11/23/2020 14:49

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: STUBBS

DATE/TIME OF TEST: 10/30/2020 22:59:02

Diagnostics Check		OK	22:50
Air Blank		0.000	22:50
Control Test		0.081	22:51
Air Blank		0.000	22:51
Subject Sample	# 1	0.090	22:54
Breath Volume		1.257	22:54
Air Blank		0.000	22:55
Air Blank		0.000	22:56
Subject Sample	# 2	0.087	22:57
Breath Volume		1.101	22:57
Air Blank		0.000	22:58
Control Test		0.078	22:58
Air Blank		0.000	22:58
Diagnostics Check		OK	22:59

Data Download Date/Time: 11/23/2020 14:49

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: GABRIEL

DATE/TIME OF TEST: 10/31/2020 02:28:58

Diagnostics Check		OK	02:15	
Air Blank		0.000	02:16	
Control Test		0.080	02:16	
Air Blank		0.000	02:17	
Subject Sample	# 1	0.149	02:20	Volume Not Met
Breath Volume		0.281	02:20	
Air Blank		0.000	02:21	
Air Blank		0.000	02:22	
Subject Sample	# 2	0.153	02:23	
Breath Volume		1.769	02:23	
Air Blank		0.000	02:24	
Air Blank		0.000	02:25	
Subject Sample	# 3	0.152	02:27	
Breath Volume		2.355	02:27	
Air Blank		0.000	02:28	
Control Test		0.079	02:28	
Air Blank		0.000	02:28	
Diagnostics Check		OK	02:28	

Data Download Date/Time: 11/23/2020 14:49

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 20:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: DREXEL JR

DATE/TIME OF TEST: 10/31/2020 21:47:17

Diagnostics Check		OK	21:31	
Air Blank		0.000	21:31	
Control Test		0.081	21:32	
Air Blank		0.000	21:32	
Subject Sample	# 1	0.211	21:35	Volume Not Met
Breath Volume		0.042	21:35	
Air Blank		0.000	21:36	
Air Blank		0.000	21:38	
Subject Sample	# 2	0.273	21:40	
Breath Volume		2.300	21:40	
Air Blank		0.000	21:41	
Air Blank		0.000	21:42	
Subject Sample	# 3	0.270	21:45	
Breath Volume		1.972	21:45	
Air Blank		0.000	21:46	
Control Test		0.076	21:46	
Air Blank		0.000	21:47	
Diagnostics Check		OK	21:47	

Data Download Date/Time: 11/23/2020 14:50

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ANDRADE,DAVID
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 10/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 21:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: ALLEN JR

DATE/TIME OF TEST: 10/31/2020 22:39:26

Diagnostics Check		OK	22:32
Air Blank		0.000	22:33
Control Test		0.079	22:33
Air Blank		0.000	22:34
Subject Sample	# 1	0.000	22:35
Breath Volume		1.855	22:35
Air Blank		0.000	22:35
Air Blank		0.000	22:37
Subject Sample	# 2	0.000	22:38
Breath Volume		2.070	22:38
Air Blank		0.000	22:38
Control Test		0.079	22:38
Air Blank		0.000	22:39
Diagnostics Check		OK	22:39

Data Download Date/Time: 11/23/2020 14:50

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.