Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

INSTRUMENT REGISTERED TO: HIALEAH GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUERTA,EDWIN,
BREATH TEST OPERATOR AGENCY: HIALEAH GARDENS PD
ARRESTING OFFICER: HUERTA,EDWIN
ARRESTING OFFICER AGENCY: HIALEAH GARDENS PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 02/15/2020 03:31:51

Diagnostics Check		OK	03:28	
Air Blank		0.000	03:29	
Control Test		0.078	03:29	
Air Blank		0.000	03:30	
Subject Sample	# 1		03:30	Subject Test Refused
Breath Volume		0.000	03:30	
Air Blank		0.000	03:30	
Control Test		0.078	03:31	
Air Blank		0.000	03:31	
Diagnostics Check		OK	03:31	

Data Download Date/Time: 02/26/2020 15:44

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

INSTRUMENT REGISTERED TO: HIALEAH GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORISMOND,WISNER,
BREATH TEST OPERATOR AGENCY: MEDLEY PD
ARRESTING OFFICER: DORISMOND,WISNER
ARRESTING OFFICER AGENCY: MEDLEY PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: SANTOS

DATE/TIME OF TEST: 02/23/2020 03:58:14

Diagnostics Check		OK	03:55	
Air Blank		0.000	03:55	
Control Test		0.078	03:56	
Air Blank		0.000	03:56	
Subject Sample	# 1		03:56	Subject Test Refused
Breath Volume		0.000	03:56	
Air Blank		0.000	03:57	
Control Test		0.078	03:57	
Air Blank		0.000	03:58	
Diagnostics Check		OK	03:58	

Data Download Date/Time: 02/26/2020 15:44

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PACHECO,GEORGE,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: PACHECO,GEORGE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 02/01/2020 02:32:58

Diagnostics Check		OK	02:25
Air Blank		0.000	02:26
Control Test		0.081	02:26
Air Blank		0.000	02:26
Subject Sample	# 1	0.193	02:27
Breath Volume		1.183	02:27
Air Blank		0.000	02:28
Air Blank		0.000	02:29
Subject Sample	# 2	0.187	02:31
Breath Volume		2.109	02:31
Air Blank		0.000	02:32
Control Test		0.080	02:32
Air Blank		0.000	02:32
Diagnostics Check		OK	02:32

Data Download Date/Time: 02/26/2020 12:58

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUBELA,CHANTEL,
BREATH TEST OPERATOR AGENCY: OTHER LOCAL
ARRESTING OFFICER: CUBELA,CHANTEL
ARRESTING OFFICER AGENCY: OTHER LOCAL
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 08:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:30

SUBJECT'S LAST NAME: PETERSON

DATE/TIME OF TEST: 02/02/2020 09:15:48

Diagnostics Check		OK	09:05	
Air Blank		0.000	09:05	
Control Test		0.081	09:06	
Air Blank		0.000	09:06	
Subject Sample	# 1		09:09	No Sample Provided
Breath Volume		0.000	09:09	
Air Blank		0.000	09:10	
Air Blank		0.000	09:12	
Subject Sample	# 2		09:14	Subject Test Refused
Breath Volume		0.000	09:14	
Air Blank		0.000	09:14	
Control Test		0.081	09:15	
Air Blank		0.000	09:15	
Diagnostics Check		OK	09:15	

Data Download Date/Time: 02/26/2020 12:58

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: DIAZ,EDWARD
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:54

SUBJECT'S LAST NAME: MORANTE

DATE/TIME OF TEST: 02/07/2020 06:20:10

Diagnostics Check		OK	06:17	
Air Blank		0.000	06:17	
Control Test		0.080	06:17	
Air Blank		0.000	06:18	
Subject Sample	# 1		06:18	Subject Test Refused
Breath Volume		0.000	06:18	
Air Blank		0.000	06:19	
Control Test		0.081	06:19	
Air Blank		0.000	06:20	
Diagnostics Check		OK	06:20	

Data Download Date/Time: 02/26/2020 12:58

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,D,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: SHIRLEY,S
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: MOLLYNEAUX

DATE/TIME OF TEST: 02/09/2020 03:11:21

Diagnostics Check		OK	03:04
Air Blank		0.000	03:04
Control Test		0.081	03:05
Air Blank		0.000	03:05
Subject Sample	# 1	0.230	03:06
Breath Volume		1.500	03:06
Air Blank		0.000	03:07
Air Blank		0.000	03:08
Subject Sample	# 2	0.224	03:09
Breath Volume		1.597	03:09
Air Blank		0.000	03:10
Control Test		0.081	03:10
Air Blank		0.000	03:11
Diagnostics Check		OK	03:11

Data Download Date/Time: 02/26/2020 12:58

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUBIN,SARAH,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: LOUIS,WALEX
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: VARGAS

DATE/TIME OF TEST: 02/09/2020 23:29:39

Diagnostics Check		OK	23:18
Air Blank		0.000	23:18
Control Test		0.080	23:19
Air Blank		0.000	23:19
Subject Sample	# 1	0.194	23:20
Breath Volume		1.957	23:20
Air Blank		0.000	23:21
Air Blank		0.000	23:22
Subject Sample	# 2	0.162	23:23
Breath Volume		1.363	23:23
Air Blank		0.000	23:24
Air Blank		0.000	23:26
Subject Sample	# 3	0.178	23:27
Breath Volume		2.175	23:27
Air Blank		0.000	23:28
Control Test		0.078	23:29
Air Blank		0.000	23:29
Diagnostics Check		OK	23:29

Data Download Date/Time: 02/26/2020 12:58

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUJILLO,GEORGE,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: TRUJILLO,GEORGE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:51

SUBJECT'S LAST NAME: MACK

DATE/TIME OF TEST: 02/14/2020 17:54:26

Diagnostics Check		OK	17:50	
Air Blank		0.000	17:51	
Control Test		0.081	17:51	
Air Blank		0.000	17:52	
Subject Sample	# 1	0.186	17:52	
Breath Volume		1.957	17:52	
Air Blank			17:53	Purge Fail
Air Blank			17:54	Purge Fail

Data Download Date/Time: 02/26/2020 12:58

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUJILLO,GEORGE,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: TRUJILLO,GEORGE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:41

SUBJECT'S LAST NAME: MACK

DATE/TIME OF TEST: 02/14/2020 18:03:13

Diagnostics Check	OK	18:01	
Air Blank	0.000	18:02	
Control Test	0.060	18:02	Control Outside Tolerance
Air Blank	0.000	18:03	

Data Download Date/Time: 02/26/2020 12:58

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUJILLO,GEORGE,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: TRUJILLO,GEORGE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:41

SUBJECT'S LAST NAME: MACK

DATE/TIME OF TEST: 02/14/2020 18:06:41

Diagnostics Check	OK	18:05	
Air Blank	0.000	18:05	
Control Test	0.069	18:06	Control Outside Tolerance
Air Blank	0.000	18:06	

Data Download Date/Time: 02/26/2020 12:58

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUJILLO,GEORGE,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: TRUJILLO,GEORGE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:41

SUBJECT'S LAST NAME: MACK

DATE/TIME OF TEST: 02/14/2020 18:09:54

Diagnostics Check	OK	18:08	
Air Blank	0.000	18:08	
Control Test	0.073	18:09	Control Outside Tolerance
Air Blank	0.000	18:09	

Data Download Date/Time: 02/26/2020 12:58

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUJILLO,GEORGE,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: TRUJILLO,GEORGE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:41

SUBJECT'S LAST NAME: MACK

DATE/TIME OF TEST: 02/14/2020 18:18:54

Diagnostics Check		OK	18:12
Air Blank		0.000	18:12
Control Test		0.075	18:13
Air Blank		0.000	18:13
Subject Sample	# 1	0.164	18:14
Breath Volume		2.003	18:14
Air Blank		0.000	18:15
Air Blank		0.000	18:16
Subject Sample	# 2	0.167	18:17
Breath Volume		1.777	18:17
Air Blank		0.000	18:17
Control Test		0.075	18:18
Air Blank		0.000	18:18
Diagnostics Check		OK	18:18

Data Download Date/Time: 02/26/2020 12:58

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUJILLO,GEORGE,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: TRUJILLO,GEORGE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:41

SUBJECT'S LAST NAME: MACK

DATE/TIME OF TEST: 02/14/2020 18:32:01

Diagnostics Check		OK	18:28	
Air Blank		0.000	18:29	
Control Test		0.078	18:29	
Air Blank		0.000	18:30	
Subject Sample	# 1	0.163	18:30	
Breath Volume		1.515	18:30	
Air Blank			18:31	Purge Fail
Air Blank			18:32	Purge Fail

Data Download Date/Time: 02/26/2020 12:58

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUJILLO,GEORGE,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: TRUJILLO,GEORGE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:41

SUBJECT'S LAST NAME: MACK

DATE/TIME OF TEST: 02/14/2020 18:36:17

Diagnostics Check	OK	18:34	
Air Blank	0.000	18:35	
Control Test	0.052	18:35	Control Outside Tolerance
Air Blank	0.000	18:36	

Data Download Date/Time: 02/26/2020 12:58

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUJILLO,GEORGE,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: TRUJILLO,GEORGE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:41

SUBJECT'S LAST NAME: MACIK

DATE/TIME OF TEST: 02/14/2020 18:39:10

Diagnostics Check	OK	18:37	
Air Blank	0.000	18:38	
Control Test	0.064	18:38	Control Outside Tolerance
Air Blank	0.000	18:39	

Data Download Date/Time: 02/26/2020 12:59

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUJILLO,GEORGE,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: TRUJILLO,GEORGE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:41

SUBJECT'S LAST NAME: MACK

DATE/TIME OF TEST: 02/14/2020 18:41:55

Diagnostics Check	OK	18:40	
Air Blank	0.000	18:40	
Control Test	0.069	18:41	Control Outside Tolerance
Air Blank	0.000	18:41	

Data Download Date/Time: 02/26/2020 12:59

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUJILLO,GEORGE,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: TRUJILLO,GEORGE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:41

SUBJECT'S LAST NAME: MACK

DATE/TIME OF TEST: 02/14/2020 18:44:40

Diagnostics Check	OK	18:43	
Air Blank	0.000	18:43	
Control Test	0.072	18:44	Control Outside Tolerance
Air Blank	0.000	18:44	

Data Download Date/Time: 02/26/2020 12:59

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUJILLO,GEORGE,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: TRUJILLO,GEORGE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:41

SUBJECT'S LAST NAME: MACK

DATE/TIME OF TEST: 02/14/2020 18:47:48

Diagnostics Check	OK	18:46	
Air Blank	0.000	18:46	
Control Test	0.073	18:47	Control Outside Tolerance
Air Blank	0.000	18:47	

Data Download Date/Time: 02/26/2020 12:59

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUJILLO,GEORGE,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: TRUJILLO,GEORGE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:41

SUBJECT'S LAST NAME: MACK

DATE/TIME OF TEST: 02/14/2020 18:58:21

Diagnostics Check	OK	18:56	
Air Blank	0.000	18:57	
Control Test	0.070	18:57	Control Outside Tolerance
Air Blank	0.000	18:58	

Data Download Date/Time: 02/26/2020 12:59

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NADAL,JOSE,V
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: NADAL,JOSE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 08:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:15

SUBJECT'S LAST NAME: KINARD

DATE/TIME OF TEST: 02/17/2020 10:38:58

Diagnostics Check		OK	10:36	
Air Blank		0.000	10:36	
Control Test		0.079	10:36	
Air Blank		0.000	10:37	
Subject Sample	# 1		10:37	Subject Test Refused
Breath Volume		0.000	10:37	
Air Blank		0.000	10:38	
Control Test		0.080	10:38	
Air Blank		0.000	10:38	
Diagnostics Check		OK	10:38	

Data Download Date/Time: 02/26/2020 12:59

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINOJOSA,DANIEL,M
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: HINOJOSA,DANIEL
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: ANAYA CALDERON

DATE/TIME OF TEST: 02/18/2020 02:53:14

Diagnostics Check		OK	02:46
Air Blank		0.000	02:46
Control Test		0.079	02:47
Air Blank		0.000	02:47
Subject Sample	# 1	0.186	02:48
Breath Volume		1.191	02:48
Air Blank		0.000	02:49
Air Blank		0.000	02:50
Subject Sample	# 2	0.188	02:51
Breath Volume		1.531	02:51
Air Blank		0.000	02:52
Control Test		0.079	02:52
Air Blank		0.000	02:53
Diagnostics Check		OK	02:53

Data Download Date/Time: 02/26/2020 12:59

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AGUILERA,ANDREW,E
BREATH TEST OPERATOR AGENCY: OTHER LOCAL
ARRESTING OFFICER: AGUILERA,ANDREW
ARRESTING OFFICER AGENCY: OTHER LOCAL
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: MORELTRINIDAD

DATE/TIME OF TEST: 02/20/2020 01:59:55

Diagnostics Check		OK	01:56	
Air Blank		0.000	01:57	
Control Test		0.080	01:57	
Air Blank		0.000	01:58	
Subject Sample	# 1		01:58	Subject Test Refused
Breath Volume		0.000	01:58	
Air Blank		0.000	01:59	
Control Test		0.080	01:59	
Air Blank		0.000	01:59	
Diagnostics Check		OK	01:59	

Data Download Date/Time: 02/26/2020 12:59

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINOJOSA,DANIEL,M
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: HINOJOSA,DANIEL
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 07:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:40

SUBJECT'S LAST NAME: CARNAVAL

DATE/TIME OF TEST: 02/20/2020 09:07:29

Diagnostics Check		OK	09:00
Air Blank		0.000	09:01
Control Test		0.080	09:01
Air Blank		0.000	09:02
Subject Sample	# 1	0.148	09:02
Breath Volume		1.996	09:02
Air Blank		0.000	09:03
Air Blank		0.000	09:05
Subject Sample	# 2	0.146	09:05
Breath Volume		2.320	09:05
Air Blank		0.000	09:06
Control Test		0.080	09:06
Air Blank		0.000	09:07
Diagnostics Check		OK	09:07

Data Download Date/Time: 02/26/2020 12:59

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,O,
BREATH TEST OPERATOR AGENCY: SWEETWATER PD
ARRESTING OFFICER: FERNANDEZ,O
ARRESTING OFFICER AGENCY: SWEETWATER PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: ORTEGA

DATE/TIME OF TEST: 02/23/2020 05:39:45

Diagnostics Check		OK	05:36	
Air Blank		0.000	05:36	
Control Test		0.081	05:37	
Air Blank		0.000	05:37	
Subject Sample	# 1		05:38	Subject Test Refused
Breath Volume		0.000	05:38	
Air Blank		0.000	05:38	
Control Test		0.080	05:39	
Air Blank		0.000	05:39	
Diagnostics Check		OK	05:39	

Data Download Date/Time: 02/26/2020 12:59

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORENZO,YOSNIEL,
BREATH TEST OPERATOR AGENCY: SWEETWATER PD
ARRESTING OFFICER: LORENZO,YOSNIEL
ARRESTING OFFICER AGENCY: SWEETWATER PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: MALAKHOV

DATE/TIME OF TEST: 02/25/2020 20:08:27

Diagnostics Check		OK	20:04	
Air Blank		0.000	20:04	
Control Test		0.081	20:04	
Air Blank		0.000	20:05	
Subject Sample	# 1		20:07	Subject Test Refused
Breath Volume		0.000	20:07	
Air Blank		0.000	20:07	
Control Test		0.081	20:07	
Air Blank		0.000	20:08	
Diagnostics Check		OK	20:08	

Data Download Date/Time: 02/26/2020 12:59

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTANA,JORGE,F
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: SHIRLEY,SAMUEL
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: FONTANIES FRANCISCO

DATE/TIME OF TEST: 02/26/2020 04:14:09

Diagnostics Check		OK	04:06
Air Blank		0.000	04:06
Control Test		0.080	04:07
Air Blank		0.000	04:07
Subject Sample	# 1	0.223	04:08
Breath Volume		1.703	04:08
Air Blank		0.000	04:09
Air Blank		0.000	04:11
Subject Sample	# 2	0.207	04:12
Breath Volume		1.703	04:12
Air Blank		0.000	04:13
Control Test		0.079	04:13
Air Blank		0.000	04:14
Diagnostics Check		OK	04:14

Data Download Date/Time: 02/26/2020 12:59

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,O,
BREATH TEST OPERATOR AGENCY: SWEETWATER PD
ARRESTING OFFICER: FERNANDEZ,O
ARRESTING OFFICER AGENCY: SWEETWATER PD
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: REDONDO LLANOS

DATE/TIME OF TEST: 02/28/2020 03:42:59

Diagnostics Check		OK	03:32
Air Blank		0.000	03:33
Control Test		0.081	03:33
Air Blank		0.000	03:33
Subject Sample	# 1	0.194	03:35
Breath Volume		4.296	03:35
Air Blank		0.000	03:36
Air Blank		0.000	03:37
Subject Sample	# 2	0.167	03:38
Breath Volume		1.363	03:38
Air Blank		0.000	03:39
Air Blank		0.000	03:40
Subject Sample	# 3	0.181	03:41
Breath Volume		4.167	03:41
Air Blank		0.000	03:42
Control Test		0.079	03:42
Air Blank		0.000	03:42
Diagnostics Check		OK	03:42

Data Download Date/Time: 03/25/2020 13:17

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,DANIEL,A
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: CRUZ,DANIEL
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:19

SUBJECT'S LAST NAME: FULLWOOD

DATE/TIME OF TEST: 02/29/2020 21:45:47

Diagnostics Check		OK	21:39
Air Blank		0.000	21:39
Control Test		0.080	21:40
Air Blank		0.000	21:40
Subject Sample	# 1	0.188	21:41
Breath Volume		1.546	21:41
Air Blank		0.000	21:42
Air Blank		0.000	21:43
Subject Sample	# 2	0.184	21:44
Breath Volume		1.839	21:44
Air Blank		0.000	21:44
Control Test		0.078	21:45
Air Blank		0.000	21:45
Diagnostics Check		OK	21:45

Data Download Date/Time: 03/25/2020 13:17

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMONS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 02/01/2020 00:58:49

Diagnostics Check		OK	00:52
Air Blank		0.000	00:52
Control Test		0.080	00:52
Air Blank		0.000	00:53
Subject Sample	# 1	0.000	00:54
Breath Volume		2.953	00:54
Air Blank		0.000	00:55
Air Blank		0.000	00:57
Subject Sample	# 2	0.000	00:57
Breath Volume		3.066	00:57
Air Blank		0.000	00:57
Control Test		0.079	00:58
Air Blank		0.000	00:58
Diagnostics Check		OK	00:58

Data Download Date/Time: 02/13/2020 09:56

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: PRENTICE

DATE/TIME OF TEST: 02/01/2020 02:08:33

Diagnostics Check		OK	02:05	
Air Blank		0.000	02:06	
Control Test		0.080	02:06	
Air Blank		0.000	02:06	
Subject Sample	# 1		02:07	Subject Test Refused
Breath Volume		0.000	02:07	
Air Blank		0.000	02:07	
Control Test		0.080	02:07	
Air Blank		0.000	02:08	
Diagnostics Check		OK	02:08	

Data Download Date/Time: 02/13/2020 09:56

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: ROSADO

DATE/TIME OF TEST: 02/01/2020 02:44:29

Diagnostics Check		OK	02:38
Air Blank		0.000	02:38
Control Test		0.080	02:39
Air Blank		0.000	02:39
Subject Sample	# 1	0.057	02:40
Breath Volume		1.957	02:40
Air Blank		0.000	02:40
Air Blank		0.000	02:42
Subject Sample	# 2	0.053	02:43
Breath Volume		1.269	02:43
Air Blank		0.000	02:43
Control Test		0.079	02:43
Air Blank		0.000	02:44
Diagnostics Check		OK	02:44

Data Download Date/Time: 02/13/2020 10:08

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 02/01/2020 03:34:00

Diagnostics Check		OK	03:22	
Air Blank		0.000	03:22	
Control Test		0.081	03:22	
Air Blank		0.000	03:23	
Subject Sample	# 1	0.191	03:26	Volume Not Met
Breath Volume		0.917	03:26	
Air Blank		0.000	03:27	
Air Blank		0.000	03:28	
Subject Sample	# 2	0.200	03:29	
Breath Volume		1.882	03:29	
Air Blank		0.000	03:30	
Air Blank		0.000	03:31	
Subject Sample	# 3	0.195	03:32	
Breath Volume		1.187	03:32	
Air Blank		0.000	03:33	
Control Test		0.078	03:33	
Air Blank		0.000	03:33	
Diagnostics Check		OK	03:34	

Data Download Date/Time: 02/13/2020 10:08

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:36

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 02/01/2020 04:59:07

Diagnostics Check		OK	04:56	
Air Blank		0.000	04:56	
Control Test		0.079	04:57	
Air Blank		0.000	04:57	
Subject Sample	# 1		04:57	Subject Test Refused
Breath Volume		0.000	04:57	
Air Blank		0.000	04:58	
Control Test		0.080	04:58	
Air Blank		0.000	04:59	
Diagnostics Check		OK	04:59	

Data Download Date/Time: 02/13/2020 10:08

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: NAVARRO-REYES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: TALLMAN

DATE/TIME OF TEST: 02/01/2020 05:57:04

Diagnostics Check		OK	05:50
Air Blank		0.000	05:51
Control Test		0.079	05:51
Air Blank		0.000	05:52
Subject Sample	# 1	0.068	05:52
Breath Volume		2.441	05:52
Air Blank		0.000	05:53
Air Blank		0.000	05:55
Subject Sample	# 2	0.070	05:55
Breath Volume		2.222	05:55
Air Blank		0.000	05:56
Control Test		0.080	05:56
Air Blank		0.000	05:56
Diagnostics Check		OK	05:57

Data Download Date/Time: 02/13/2020 10:08

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: HAMILTON

DATE/TIME OF TEST: 02/02/2020 03:34:56

Diagnostics Check		OK	03:28
Air Blank		0.000	03:29
Control Test		0.080	03:29
Air Blank		0.000	03:29
Subject Sample	# 1	0.148	03:30
Breath Volume		2.648	03:30
Air Blank		0.000	03:30
Air Blank		0.000	03:32
Subject Sample	# 2	0.139	03:33
Breath Volume		1.906	03:33
Air Blank		0.000	03:34
Control Test		0.080	03:34
Air Blank		0.000	03:34
Diagnostics Check		OK	03:34

Data Download Date/Time: 02/13/2020 10:08

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: PRAY

DATE/TIME OF TEST: 02/03/2020 02:45:46

Diagnostics Check		OK	02:41	
Air Blank		0.000	02:41	
Control Test		0.080	02:42	
Air Blank		0.000	02:42	
Subject Sample	# 1		02:44	Subject Test Refused
Breath Volume		0.000	02:44	
Air Blank		0.000	02:44	
Control Test		0.080	02:45	
Air Blank		0.000	02:45	
Diagnostics Check		OK	02:45	

Data Download Date/Time: 02/13/2020 10:08

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HANSON,T
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: ARMSTRONG

DATE/TIME OF TEST: 02/03/2020 04:21:57

Diagnostics Check		OK	04:19	
Air Blank		0.000	04:19	
Control Test		0.081	04:19	
Air Blank		0.000	04:20	
Subject Sample	# 1		04:20	Subject Test Refused
Breath Volume		0.000	04:20	
Air Blank		0.000	04:21	
Control Test		0.080	04:21	
Air Blank		0.000	04:21	
Diagnostics Check		OK	04:21	

Data Download Date/Time: 02/13/2020 10:08

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAWSON,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 02/04/2020 01:52:29

Diagnostics Check		OK	01:47	
Air Blank		0.000	01:48	
Control Test		0.080	01:48	
Air Blank		0.000	01:49	
Subject Sample	# 1		01:51	Subject Test Refused
Breath Volume		0.035	01:51	
Air Blank		0.000	01:51	
Control Test		0.080	01:51	
Air Blank		0.000	01:52	
Diagnostics Check		OK	01:52	

Data Download Date/Time: 02/13/2020 10:08

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,M,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: NORRIS

DATE/TIME OF TEST: 02/07/2020 04:05:22

Diagnostics Check		OK	04:02	
Air Blank		0.000	04:03	
Control Test		0.080	04:03	
Air Blank		0.000	04:03	
Subject Sample	# 1		04:04	Subject Test Refused
Breath Volume		0.000	04:04	
Air Blank		0.000	04:04	
Control Test		0.080	04:04	
Air Blank		0.000	04:05	
Diagnostics Check		OK	04:05	

Data Download Date/Time: 02/13/2020 10:08

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,M,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:02

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 02/07/2020 20:28:46

Diagnostics Check		OK	20:22
Air Blank		0.000	20:23
Control Test		0.081	20:23
Air Blank		0.000	20:23
Subject Sample	# 1	0.112	20:24
Breath Volume		1.667	20:24
Air Blank		0.000	20:24
Air Blank		0.000	20:26
Subject Sample	# 2	0.108	20:27
Breath Volume		1.597	20:27
Air Blank		0.000	20:27
Control Test		0.080	20:28
Air Blank		0.000	20:28
Diagnostics Check		OK	20:28

Data Download Date/Time: 02/13/2020 10:08

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,M,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VANHOF,J
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: WADDELL

DATE/TIME OF TEST: 02/08/2020 04:09:30

Diagnostics Check		OK	04:03
Air Blank		0.000	04:03
Control Test		0.081	04:04
Air Blank		0.000	04:04
Subject Sample	# 1	0.125	04:05
Breath Volume		1.578	04:05
Air Blank		0.000	04:05
Air Blank		0.000	04:07
Subject Sample	# 2	0.122	04:08
Breath Volume		1.296	04:08
Air Blank		0.000	04:08
Control Test		0.080	04:08
Air Blank		0.000	04:09
Diagnostics Check		OK	04:09

Data Download Date/Time: 02/13/2020 10:08

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CAMACHO,K
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:39

SUBJECT'S LAST NAME: BARKER

DATE/TIME OF TEST: 02/08/2020 19:02:17

Diagnostics Check		OK	18:59	
Air Blank		0.000	18:59	
Control Test		0.080	19:00	
Air Blank		0.000	19:00	
Subject Sample	# 1		19:00	Subject Test Refused
Breath Volume		0.000	19:00	
Air Blank		0.000	19:01	
Control Test		0.080	19:01	
Air Blank		0.000	19:02	
Diagnostics Check		OK	19:02	

Data Download Date/Time: 02/13/2020 10:08

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,M,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: MARTZ

DATE/TIME OF TEST: 02/08/2020 22:20:31

Diagnostics Check		OK	22:14
Air Blank		0.000	22:14
Control Test		0.079	22:15
Air Blank		0.000	22:15
Subject Sample	# 1	0.000	22:16
Breath Volume		1.531	22:16
Air Blank		0.000	22:16
Air Blank		0.000	22:18
Subject Sample	# 2	0.000	22:19
Breath Volume		2.082	22:19
Air Blank		0.000	22:19
Control Test		0.080	22:19
Air Blank		0.000	22:20
Diagnostics Check		OK	22:20

Data Download Date/Time: 02/13/2020 10:08

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAWSON,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:59

SUBJECT'S LAST NAME: WOODY

DATE/TIME OF TEST: 02/09/2020 21:25:17

Diagnostics Check		OK	21:19
Air Blank		0.000	21:20
Control Test		0.080	21:20
Air Blank		0.000	21:20
Subject Sample	# 1	0.000	21:21
Breath Volume		1.878	21:21
Air Blank		0.000	21:21
Air Blank		0.000	21:23
Subject Sample	# 2	0.000	21:23
Breath Volume		2.230	21:23
Air Blank		0.000	21:24
Control Test		0.080	21:24
Air Blank		0.000	21:25
Diagnostics Check		OK	21:25

Data Download Date/Time: 02/13/2020 10:09

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,M,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: BURNEY

DATE/TIME OF TEST: 02/09/2020 22:30:46

Diagnostics Check		OK	22:24
Air Blank		0.000	22:25
Control Test		0.080	22:25
Air Blank		0.000	22:25
Subject Sample	# 1	0.180	22:26
Breath Volume		1.531	22:26
Air Blank		0.000	22:26
Air Blank		0.000	22:28
Subject Sample	# 2	0.174	22:29
Breath Volume		1.214	22:29
Air Blank		0.000	22:29
Control Test		0.080	22:30
Air Blank		0.000	22:30
Diagnostics Check		OK	22:30

Data Download Date/Time: 02/13/2020 10:09

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRANCH,MORGAN,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: ZENTENO

DATE/TIME OF TEST: 02/10/2020 01:12:02

Diagnostics Check		OK	01:05
Air Blank		0.000	01:05
Control Test		0.080	01:06
Air Blank		0.000	01:06
Subject Sample	# 1	0.132	01:07
Breath Volume		1.250	01:07
Air Blank		0.000	01:07
Air Blank		0.000	01:09
Subject Sample	# 2	0.117	01:10
Breath Volume		1.250	01:10
Air Blank		0.000	01:11
Control Test		0.080	01:11
Air Blank		0.000	01:11
Diagnostics Check		OK	01:12

Data Download Date/Time: 02/13/2020 10:09

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROSBY,CHRISTIAN,B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 02/11/2020 01:10:17

Diagnostics Check		OK	01:03
Air Blank		0.000	01:03
Control Test		0.079	01:03
Air Blank		0.000	01:04
Subject Sample	# 1	0.229	01:05
Breath Volume		2.355	01:05
Air Blank		0.000	01:06
Air Blank		0.000	01:08
Subject Sample	# 2	0.233	01:08
Breath Volume		2.562	01:08
Air Blank		0.000	01:09
Control Test		0.079	01:09
Air Blank		0.000	01:10
Diagnostics Check		OK	01:10

Data Download Date/Time: 02/13/2020 10:09

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROHDE,A,G
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: FRED

DATE/TIME OF TEST: 02/11/2020 04:10:48

Diagnostics Check		OK	04:08	
Air Blank		0.000	04:08	
Control Test		0.080	04:08	
Air Blank		0.000	04:09	
Subject Sample	# 1		04:09	Subject Test Refused
Breath Volume		0.000	04:09	
Air Blank		0.000	04:09	
Control Test		0.079	04:10	
Air Blank		0.000	04:10	
Diagnostics Check		OK	04:10	

Data Download Date/Time: 02/13/2020 10:09

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 02/11/2020 23:45:06

Diagnostics Check		OK	23:42	
Air Blank		0.000	23:42	
Control Test		0.080	23:42	
Air Blank		0.000	23:43	
Subject Sample	# 1		23:43	Subject Test Refused
Breath Volume		0.000	23:43	
Air Blank		0.000	23:44	
Control Test		0.080	23:44	
Air Blank		0.000	23:45	
Diagnostics Check		OK	23:45	

Data Download Date/Time: 02/13/2020 10:09

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: RAMIREZ-RAMIREZ

DATE/TIME OF TEST: 02/12/2020 02:31:04

Diagnostics Check		OK	02:24
Air Blank		0.000	02:24
Control Test		0.080	02:24
Air Blank		0.000	02:25
Subject Sample	# 1	0.163	02:26
Breath Volume		2.421	02:26
Air Blank		0.000	02:26
Air Blank		0.000	02:28
Subject Sample	# 2	0.167	02:29
Breath Volume		1.597	02:29
Air Blank		0.000	02:30
Control Test		0.080	02:30
Air Blank		0.000	02:30
Diagnostics Check		OK	02:31

Data Download Date/Time: 02/13/2020 10:09

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:27

SUBJECT'S LAST NAME: MUSSER

DATE/TIME OF TEST: 02/12/2020 20:56:43

Diagnostics Check		OK	20:49
Air Blank		0.000	20:50
Control Test		0.079	20:50
Air Blank		0.000	20:51
Subject Sample	# 1	0.017	20:52
Breath Volume		2.214	20:52
Air Blank		0.000	20:52
Air Blank		0.000	20:54
Subject Sample	# 2	0.017	20:55
Breath Volume		2.285	20:55
Air Blank		0.000	20:55
Control Test		0.080	20:56
Air Blank		0.000	20:56
Diagnostics Check		OK	20:56

Data Download Date/Time: 02/13/2020 10:09

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: ACUFF

DATE/TIME OF TEST: 02/12/2020 23:43:02

Diagnostics Check		OK	23:40	
Air Blank		0.000	23:40	
Control Test		0.080	23:40	
Air Blank		0.000	23:41	
Subject Sample	# 1		23:41	Subject Test Refused
Breath Volume		0.000	23:41	
Air Blank		0.000	23:42	
Control Test		0.080	23:42	
Air Blank		0.000	23:42	
Diagnostics Check		OK	23:43	

Data Download Date/Time: 02/13/2020 10:09

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,G
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 02/13/2020 01:42:23

Diagnostics Check		OK	01:31	
Air Blank		0.000	01:31	
Control Test		0.079	01:31	
Air Blank		0.000	01:32	
Subject Sample	# 1		01:35	No Sample Provided
Breath Volume		0.000	01:35	
Air Blank		0.000	01:35	
Air Blank		0.000	01:37	
Subject Sample	# 2		01:41	No Sample Provided
Breath Volume		0.000	01:41	
Air Blank		0.000	01:41	
Control Test		0.080	01:41	
Air Blank		0.000	01:42	
Diagnostics Check		OK	01:42	

Data Download Date/Time: 02/13/2020 10:09

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ELEFTHERIADIS,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: CHAPMAN

DATE/TIME OF TEST: 02/13/2020 04:48:47

Diagnostics Check		OK	04:32	
Air Blank		0.000	04:33	
Control Test		0.079	04:33	
Air Blank		0.000	04:33	
Subject Sample	# 1	0.114	04:37	Volume Not Met
Breath Volume		0.277	04:37	
Air Blank		0.000	04:37	
Air Blank		0.000	04:39	
Subject Sample	# 2	0.136	04:42	
Breath Volume		1.484	04:42	
Air Blank		0.000	04:43	
Air Blank		0.000	04:44	
Subject Sample	# 3	0.134	04:47	
Breath Volume		1.273	04:47	
Air Blank		0.000	04:47	
Control Test		0.078	04:48	
Air Blank		0.000	04:48	
Diagnostics Check		OK	04:48	

Data Download Date/Time: 02/13/2020 10:09

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOVACS,TYLER,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 02/14/2020 23:03:05

Diagnostics Check		OK	22:55
Air Blank		0.000	22:56
Control Test		0.080	22:56
Air Blank		0.000	22:57
Subject Sample	# 1	0.028	22:57
Breath Volume		1.300	22:57
Air Blank		0.000	22:58
Air Blank		0.000	23:00
Subject Sample	# 2	0.025	23:01
Breath Volume		2.343	23:01
Air Blank		0.000	23:02
Control Test		0.080	23:02
Air Blank		0.000	23:03
Diagnostics Check		OK	23:03

Data Download Date/Time: 03/03/2020 14:34

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: ELLGID

DATE/TIME OF TEST: 02/15/2020 00:44:17

Diagnostics Check		OK	00:37
Air Blank		0.000	00:38
Control Test		0.081	00:38
Air Blank		0.000	00:39
Subject Sample	# 1	0.173	00:39
Breath Volume		2.191	00:39
Air Blank		0.000	00:40
Air Blank		0.000	00:42
Subject Sample	# 2	0.171	00:42
Breath Volume		2.000	00:42
Air Blank		0.000	00:43
Control Test		0.080	00:43
Air Blank		0.000	00:44
Diagnostics Check		OK	00:44

Data Download Date/Time: 03/03/2020 14:34

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOVACS,TYLER,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: CLARKE

DATE/TIME OF TEST: 02/15/2020 02:09:45

Diagnostics Check		OK	02:01
Air Blank		0.000	02:01
Control Test		0.080	02:02
Air Blank		0.000	02:02
Subject Sample	# 1	0.189	02:05
Breath Volume		2.449	02:05
Air Blank		0.000	02:05
Air Blank		0.000	02:07
Subject Sample	# 2	0.189	02:08
Breath Volume		2.171	02:08
Air Blank		0.000	02:08
Control Test		0.079	02:09
Air Blank		0.000	02:09
Diagnostics Check		OK	02:09

Data Download Date/Time: 03/03/2020 14:34

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOVACS,TYLER,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMONS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:24

SUBJECT'S LAST NAME: WERRE

DATE/TIME OF TEST: 02/15/2020 21:47:39

Diagnostics Check		OK	21:44	
Air Blank		0.000	21:45	
Control Test		0.080	21:45	
Air Blank		0.000	21:46	
Subject Sample	# 1		21:46	Subject Test Refused
Breath Volume		0.000	21:46	
Air Blank		0.000	21:46	
Control Test		0.080	21:47	
Air Blank		0.000	21:47	
Diagnostics Check		OK	21:47	

Data Download Date/Time: 03/03/2020 14:34

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOVACS,TYLER,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMONS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: KOSOVA

DATE/TIME OF TEST: 02/16/2020 04:51:33

Diagnostics Check		OK	04:38	
Air Blank		0.000	04:38	
Control Test		0.080	04:39	
Air Blank		0.000	04:39	
Subject Sample	# 1	0.212	04:41	
Breath Volume		2.257	04:41	
Air Blank		0.000	04:41	
Air Blank		0.000	04:43	
Subject Sample	# 2	0.206	04:46	Volume Not Met
Breath Volume		0.042	04:46	
Air Blank		0.000	04:47	
Air Blank		0.000	04:49	
Subject Sample	# 3	0.198	04:50	
Breath Volume		1.246	04:50	
Air Blank		0.000	04:50	
Control Test		0.077	04:50	
Air Blank		0.000	04:51	
Diagnostics Check		OK	04:51	

Data Download Date/Time: 03/03/2020 14:34

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOVACS,TYLER,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CLARK,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: FRANKOSKI

DATE/TIME OF TEST: 02/17/2020 00:37:06

Diagnostics Check		OK	00:34	
Air Blank		0.000	00:34	
Control Test		0.079	00:34	
Air Blank		0.000	00:35	
Subject Sample	# 1		00:35	Subject Test Refused
Breath Volume		0.000	00:35	
Air Blank		0.000	00:36	
Control Test		0.080	00:36	
Air Blank		0.000	00:37	
Diagnostics Check		OK	00:37	

Data Download Date/Time: 03/03/2020 14:35

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISHER,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:58

SUBJECT'S LAST NAME: HOLMES

DATE/TIME OF TEST: 02/17/2020 20:25:03

Diagnostics Check		OK	20:19
Air Blank		0.000	20:19
Control Test		0.080	20:19
Air Blank		0.000	20:20
Subject Sample	# 1	0.076	20:20
Breath Volume		2.597	20:20
Air Blank		0.000	20:21
Air Blank		0.000	20:23
Subject Sample	# 2	0.075	20:23
Breath Volume		2.957	20:23
Air Blank		0.000	20:24
Control Test		0.080	20:24
Air Blank		0.000	20:24
Diagnostics Check		OK	20:25

Data Download Date/Time: 03/03/2020 14:35

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOVACS,TYLER,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: NUNLEY

DATE/TIME OF TEST: 02/21/2020 00:50:04

Diagnostics Check		OK	00:43
Air Blank		0.000	00:44
Control Test		0.080	00:44
Air Blank		0.000	00:45
Subject Sample	# 1	0.000	00:45
Breath Volume		1.531	00:45
Air Blank		0.000	00:46
Air Blank		0.000	00:48
Subject Sample	# 2	0.000	00:48
Breath Volume		3.785	00:48
Air Blank		0.000	00:49
Control Test		0.081	00:49
Air Blank		0.000	00:49
Diagnostics Check		OK	00:50

Data Download Date/Time: 03/03/2020 14:35

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOVACS,TYLER,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: VEASLEY

DATE/TIME OF TEST: 02/22/2020 00:26:50

Diagnostics Check		OK	00:24	
Air Blank		0.000	00:24	
Control Test		0.081	00:24	
Air Blank		0.000	00:25	
Subject Sample	# 1		00:26	RFI Detect
Breath Volume		0.644	00:26	
Air Blank			00:26	RFI Detect
Air Blank		0.000	00:26	

Data Download Date/Time: 03/03/2020 14:35

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOVACS,TYLER,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/13/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: VEASLEY

DATE/TIME OF TEST: 02/22/2020 00:36:20

Diagnostics Check		OK	00:29
Air Blank		0.000	00:30
Control Test		0.081	00:30
Air Blank		0.000	00:31
Subject Sample	# 1	0.143	00:31
Breath Volume		1.675	00:31
Air Blank		0.000	00:32
Air Blank		0.000	00:34
Subject Sample	# 2	0.141	00:34
Breath Volume		2.191	00:34
Air Blank		0.000	00:35
Control Test		0.080	00:35
Air Blank		0.000	00:36
Diagnostics Check		OK	00:36

Data Download Date/Time: 03/03/2020 14:35

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOVACS,TYLER,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SHEA,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: WHITLEY

DATE/TIME OF TEST: 02/22/2020 02:24:33

Diagnostics Check		OK	02:18
Air Blank		0.000	02:18
Control Test		0.081	02:19
Air Blank		0.000	02:19
Subject Sample	# 1	0.000	02:20
Breath Volume		2.136	02:20
Air Blank		0.000	02:20
Air Blank		0.000	02:22
Subject Sample	# 2	0.000	02:23
Breath Volume		2.507	02:23
Air Blank		0.000	02:23
Control Test		0.081	02:23
Air Blank		0.000	02:24
Diagnostics Check		OK	02:24

Data Download Date/Time: 03/03/2020 14:35

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOVACS,TYLER,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:29

SUBJECT'S LAST NAME: BATTLE

DATE/TIME OF TEST: 02/22/2020 20:52:48

Diagnostics Check		OK	20:49	
Air Blank		0.000	20:50	
Control Test		0.079	20:50	
Air Blank		0.000	20:51	
Subject Sample	# 1		20:51	Subject Test Refused
Breath Volume		0.000	20:51	
Air Blank		0.000	20:51	
Control Test		0.080	20:52	
Air Blank		0.000	20:52	
Diagnostics Check		OK	20:52	

Data Download Date/Time: 03/03/2020 14:35

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOVACS,TYLER,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: ORS

DATE/TIME OF TEST: 02/22/2020 21:51:59

Diagnostics Check		OK	21:45
Air Blank		0.000	21:45
Control Test		0.081	21:46
Air Blank		0.000	21:46
Subject Sample	# 1	0.075	21:47
Breath Volume		2.621	21:47
Air Blank		0.000	21:47
Air Blank		0.000	21:49
Subject Sample	# 2	0.067	21:50
Breath Volume		2.546	21:50
Air Blank		0.000	21:51
Control Test		0.080	21:51
Air Blank		0.000	21:51
Diagnostics Check		OK	21:51

Data Download Date/Time: 03/03/2020 14:35

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOVACS,TYLER,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAWSON,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: CRUSE

DATE/TIME OF TEST: 02/23/2020 00:36:41

Diagnostics Check		OK	00:29
Air Blank		0.000	00:29
Control Test		0.080	00:30
Air Blank		0.000	00:30
Subject Sample	# 1	0.178	00:31
Breath Volume		1.570	00:31
Air Blank		0.000	00:32
Air Blank		0.000	00:34
Subject Sample	# 2	0.168	00:35
Breath Volume		2.097	00:35
Air Blank		0.000	00:35
Control Test		0.080	00:36
Air Blank		0.000	00:36
Diagnostics Check		OK	00:36

Data Download Date/Time: 03/03/2020 14:35

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOVACS,TYLER,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TORRES,W
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 02/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 02/23/2020 04:32:32

Diagnostics Check		OK	04:28	
Air Blank		0.000	04:28	
Control Test		0.078	04:28	
Air Blank		0.000	04:29	
Subject Sample	# 1		04:31	Subject Test Refused
Breath Volume		0.000	04:31	
Air Blank		0.000	04:31	
Control Test		0.080	04:31	
Air Blank		0.000	04:32	
Diagnostics Check		OK	04:32	

Data Download Date/Time: 03/03/2020 14:35

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAWSON,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: OLSEN

DATE/TIME OF TEST: 02/21/2020 00:34:53

Diagnostics Check		OK	00:28
Air Blank		0.000	00:29
Control Test		0.079	00:29
Air Blank		0.000	00:29
Subject Sample	# 1	0.242	00:30
Breath Volume		1.867	00:30
Air Blank		0.000	00:30
Air Blank		0.000	00:32
Subject Sample	# 2	0.235	00:33
Breath Volume		1.980	00:33
Air Blank		0.000	00:33
Control Test		0.079	00:34
Air Blank		0.000	00:34
Diagnostics Check		OK	00:34

Data Download Date/Time: 02/28/2020 13:02

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: WELSS

DATE/TIME OF TEST: 02/21/2020 23:23:35

Diagnostics Check		OK	23:17
Air Blank		0.000	23:17
Control Test		0.080	23:17
Air Blank		0.000	23:18
Subject Sample	# 1	0.120	23:19
Breath Volume		2.257	23:19
Air Blank		0.000	23:19
Air Blank		0.000	23:21
Subject Sample	# 2	0.114	23:22
Breath Volume		1.285	23:22
Air Blank		0.000	23:22
Control Test		0.079	23:22
Air Blank		0.000	23:23
Diagnostics Check		OK	23:23

Data Download Date/Time: 02/28/2020 13:02

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: FIECCONI

DATE/TIME OF TEST: 02/22/2020 01:28:22

Diagnostics Check		OK	01:25	
Air Blank		0.000	01:26	
Control Test		0.080	01:26	
Air Blank		0.000	01:26	
Subject Sample	# 1		01:27	Subject Test Refused
Breath Volume		0.000	01:27	
Air Blank		0.000	01:27	
Control Test		0.080	01:27	
Air Blank		0.000	01:28	
Diagnostics Check		OK	01:28	

Data Download Date/Time: 02/28/2020 13:02

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: BRENTS

DATE/TIME OF TEST: 02/22/2020 03:18:36

Diagnostics Check		OK	03:11
Air Blank		0.000	03:12
Control Test		0.080	03:12
Air Blank		0.000	03:13
Subject Sample	# 1	0.109	03:14
Breath Volume		1.191	03:14
Air Blank		0.000	03:14
Air Blank		0.000	03:16
Subject Sample	# 2	0.109	03:17
Breath Volume		1.125	03:17
Air Blank		0.000	03:17
Control Test		0.079	03:18
Air Blank		0.000	03:18
Diagnostics Check		OK	03:18

Data Download Date/Time: 02/28/2020 13:02

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:44

SUBJECT'S LAST NAME: BOYD

DATE/TIME OF TEST: 02/22/2020 20:14:28

Diagnostics Check		OK	20:08
Air Blank		0.000	20:08
Control Test		0.079	20:08
Air Blank		0.000	20:09
Subject Sample	# 1	0.138	20:10
Breath Volume		2.242	20:10
Air Blank		0.000	20:10
Air Blank		0.000	20:12
Subject Sample	# 2	0.126	20:12
Breath Volume		1.246	20:12
Air Blank		0.000	20:13
Control Test		0.079	20:13
Air Blank		0.000	20:14
Diagnostics Check		OK	20:14

Data Download Date/Time: 02/28/2020 13:02

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: JIMENEZ

DATE/TIME OF TEST: 02/23/2020 00:52:17

Diagnostics Check		OK	00:49	
Air Blank		0.000	00:49	
Control Test		0.079	00:50	
Air Blank		0.000	00:50	
Subject Sample	# 1		00:50	Subject Test Refused
Breath Volume		0.000	00:50	
Air Blank		0.000	00:51	
Control Test		0.078	00:51	
Air Blank		0.000	00:52	
Diagnostics Check		OK	00:52	

Data Download Date/Time: 02/28/2020 13:02

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: BAUER

DATE/TIME OF TEST: 02/23/2020 04:21:08

Diagnostics Check		OK	04:18	
Air Blank		0.000	04:18	
Control Test		0.078	04:19	
Air Blank		0.000	04:19	
Subject Sample	# 1		04:19	Subject Test Refused
Breath Volume		0.000	04:19	
Air Blank		0.000	04:20	
Control Test		0.078	04:20	
Air Blank		0.000	04:21	
Diagnostics Check		OK	04:21	

Data Download Date/Time: 02/28/2020 13:02

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 02/02/2020 03:13:34

Diagnostics Check		OK	03:10	
Air Blank		0.000	03:11	
Control Test		0.080	03:11	
Air Blank		0.000	03:12	
Subject Sample	# 1		03:12	Subject Test Refused
Breath Volume		0.000	03:12	
Air Blank		0.000	03:12	
Control Test		0.081	03:12	
Air Blank		0.000	03:13	
Diagnostics Check		OK	03:13	

Data Download Date/Time: 02/12/2020 14:28

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: MCDERMOTT

DATE/TIME OF TEST: 02/03/2020 01:36:14

Diagnostics Check		OK	01:33	
Air Blank		0.000	01:33	
Control Test		0.080	01:34	
Air Blank		0.000	01:34	
Subject Sample	# 1		01:34	Subject Test Refused
Breath Volume		0.000	01:34	
Air Blank		0.000	01:35	
Control Test		0.080	01:35	
Air Blank		0.000	01:36	
Diagnostics Check		OK	01:36	

Data Download Date/Time: 02/12/2020 14:28

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: LEWIS JR

DATE/TIME OF TEST: 02/03/2020 02:40:11

Diagnostics Check		OK	02:37	
Air Blank		0.000	02:37	
Control Test		0.080	02:38	
Air Blank		0.000	02:38	
Subject Sample	# 1		02:38	Subject Test Refused
Breath Volume		0.000	02:38	
Air Blank		0.000	02:39	
Control Test		0.080	02:39	
Air Blank		0.000	02:40	
Diagnostics Check		OK	02:40	

Data Download Date/Time: 02/12/2020 14:28

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: GRANO

DATE/TIME OF TEST: 02/03/2020 02:53:42

Diagnostics Check		OK	02:50	
Air Blank		0.000	02:51	
Control Test		0.081	02:51	
Air Blank		0.000	02:52	
Subject Sample	# 1		02:52	Subject Test Refused
Breath Volume		0.000	02:52	
Air Blank		0.000	02:52	
Control Test		0.080	02:53	
Air Blank		0.000	02:53	
Diagnostics Check		OK	02:53	

Data Download Date/Time: 02/12/2020 14:28

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CARLSON,P
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: JARED

DATE/TIME OF TEST: 02/03/2020 04:17:13

Diagnostics Check		OK	04:14	
Air Blank		0.000	04:14	
Control Test		0.080	04:14	
Air Blank		0.000	04:15	
Subject Sample	# 1		04:15	Subject Test Refused
Breath Volume		0.000	04:15	
Air Blank		0.000	04:16	
Control Test		0.081	04:16	
Air Blank		0.000	04:17	
Diagnostics Check		OK	04:17	

Data Download Date/Time: 02/12/2020 14:28

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: PATTON

DATE/TIME OF TEST: 02/04/2020 00:28:27

Diagnostics Check		OK	00:20
Air Blank		0.000	00:21
Control Test		0.080	00:21
Air Blank		0.000	00:22
Subject Sample	# 1	0.227	00:23
Breath Volume		2.445	00:23
Air Blank		0.000	00:24
Air Blank		0.000	00:26
Subject Sample	# 2	0.220	00:26
Breath Volume		2.019	00:26
Air Blank		0.000	00:27
Control Test		0.078	00:27
Air Blank		0.000	00:28
Diagnostics Check		OK	00:28

Data Download Date/Time: 02/12/2020 14:28

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: FOWLER

DATE/TIME OF TEST: 02/04/2020 02:24:50

Diagnostics Check		OK	02:18
Air Blank		0.000	02:18
Control Test		0.079	02:18
Air Blank		0.000	02:19
Subject Sample	# 1	0.096	02:20
Breath Volume		1.722	02:20
Air Blank		0.000	02:21
Air Blank		0.000	02:22
Subject Sample	# 2	0.091	02:23
Breath Volume		1.761	02:23
Air Blank		0.000	02:23
Control Test		0.078	02:24
Air Blank		0.000	02:24
Diagnostics Check		OK	02:24

Data Download Date/Time: 02/12/2020 14:28

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: SCHWARTZ

DATE/TIME OF TEST: 02/06/2020 01:25:27

Diagnostics Check		OK	01:18
Air Blank		0.000	01:18
Control Test		0.079	01:19
Air Blank		0.000	01:19
Subject Sample	# 1	0.176	01:20
Breath Volume		3.085	01:20
Air Blank		0.000	01:21
Air Blank		0.000	01:23
Subject Sample	# 2	0.169	01:23
Breath Volume		2.835	01:23
Air Blank		0.000	01:24
Control Test		0.079	01:24
Air Blank		0.000	01:25
Diagnostics Check		OK	01:25

Data Download Date/Time: 02/12/2020 14:28

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANFORD,BRIAN,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMONS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 02/09/2020 02:59:40

Diagnostics Check		OK	02:53
Air Blank		0.000	02:53
Control Test		0.080	02:53
Air Blank		0.000	02:54
Subject Sample	# 1	0.218	02:54
Breath Volume		3.304	02:54
Air Blank		0.000	02:55
Air Blank		0.000	02:57
Subject Sample	# 2	0.211	02:58
Breath Volume		2.433	02:58
Air Blank		0.000	02:58
Control Test		0.079	02:59
Air Blank		0.000	02:59
Diagnostics Check		OK	02:59

Data Download Date/Time: 02/12/2020 14:28

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001285

INSTRUMENT REGISTERED TO: GILCHRIST COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,RICK,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BROWN,RICK
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1048227
DRY GAS EXPIRY DATE: 11/15/2021
VIOLATION CODE: DUI
ARREST TIME: 16:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:57

SUBJECT'S LAST NAME: ROTHMAN

DATE/TIME OF TEST: 02/12/2020 17:25:18

Diagnostics Check		OK	17:19	
Air Blank		0.000	17:20	
Control Test		0.080	17:20	
Air Blank		0.000	17:21	
Subject Sample	# 1	0.211	17:21	
Breath Volume		1.312	17:21	
Air Blank		0.000	17:22	
Air Blank			17:23	RFI Detect
Air Blank		0.000	17:24	
Control Test		0.080	17:24	
Air Blank		0.000	17:25	
Diagnostics Check		OK	17:25	

Data Download Date/Time: 03/10/2020 12:31

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001285

INSTRUMENT REGISTERED TO: GILCHRIST COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,RICK,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BROWN,RICK
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1048227
DRY GAS EXPIRY DATE: 11/15/2021
VIOLATION CODE: DUI
ARREST TIME: 16:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:57

SUBJECT'S LAST NAME: ROTHMAN

DATE/TIME OF TEST: 02/12/2020 17:32:01

Diagnostics Check		OK	17:28	
Air Blank		0.000	17:29	
Control Test		0.079	17:29	
Air Blank		0.000	17:29	
Subject Sample	# 1	0.206	17:30	
Breath Volume		1.953	17:30	
Air Blank			17:30	RFI Detect
Air Blank		0.000	17:31	
Control Test		0.080	17:31	
Air Blank		0.000	17:31	
Diagnostics Check		OK	17:32	

Data Download Date/Time: 03/10/2020 12:31

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001285

INSTRUMENT REGISTERED TO: GILCHRIST COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,RICK,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BROWN,RICL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1048227
DRY GAS EXPIRY DATE: 11/15/2021
VIOLATION CODE: DUI
ARREST TIME: 16:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:57

SUBJECT'S LAST NAME: ROTHMAN

DATE/TIME OF TEST: 02/12/2020 17:42:17

Diagnostics Check		OK	17:36
Air Blank		0.000	17:36
Control Test		0.079	17:36
Air Blank		0.000	17:37
Subject Sample	# 1	0.205	17:37
Breath Volume		2.515	17:37
Air Blank		0.000	17:38
Air Blank		0.000	17:40
Subject Sample	# 2	0.211	17:40
Breath Volume		2.828	17:40
Air Blank		0.000	17:41
Control Test		0.077	17:41
Air Blank		0.000	17:42
Diagnostics Check		OK	17:42

Data Download Date/Time: 03/10/2020 12:31

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

INSTRUMENT REGISTERED TO: BAKER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CJ,
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MILLER,CJ
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:46

SUBJECT'S LAST NAME: BOSTON

DATE/TIME OF TEST: 02/08/2020 18:13:57

Diagnostics Check		OK	18:06
Air Blank		0.000	18:07
Control Test		0.080	18:07
Air Blank		0.000	18:08
Subject Sample	# 1	0.144	18:09
Breath Volume		2.335	18:09
Air Blank		0.000	18:09
Air Blank		0.000	18:11
Subject Sample	# 2	0.145	18:12
Breath Volume		2.796	18:12
Air Blank		0.000	18:13
Control Test		0.080	18:13
Air Blank		0.000	18:13
Diagnostics Check		OK	18:13

Data Download Date/Time: 02/28/2020 16:03

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

INSTRUMENT REGISTERED TO: BAKER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOFMANN,TRAVIS,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HOFMANN,TRAVIS
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: MCCUNE

DATE/TIME OF TEST: 02/14/2020 22:53:23

Diagnostics Check		OK	22:47
Air Blank		0.000	22:47
Control Test		0.081	22:48
Air Blank		0.000	22:48
Subject Sample	# 1	0.000	22:49
Breath Volume		1.679	22:49
Air Blank		0.000	22:49
Air Blank		0.000	22:51
Subject Sample	# 2	0.000	22:52
Breath Volume		2.171	22:52
Air Blank		0.000	22:52
Control Test		0.081	22:52
Air Blank		0.000	22:53
Diagnostics Check		OK	22:53

Data Download Date/Time: 02/28/2020 16:03

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

INSTRUMENT REGISTERED TO: BAKER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAEYAERT,JEFF,H
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: NAEYAERT,JEFF
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: ARANDA

DATE/TIME OF TEST: 02/23/2020 01:55:44

Diagnostics Check		OK	01:49
Air Blank		0.000	01:50
Control Test		0.080	01:50
Air Blank		0.000	01:50
Subject Sample	# 1	0.172	01:51
Breath Volume		2.410	01:51
Air Blank		0.000	01:51
Air Blank		0.000	01:53
Subject Sample	# 2	0.169	01:54
Breath Volume		2.621	01:54
Air Blank		0.000	01:54
Control Test		0.079	01:55
Air Blank		0.000	01:55
Diagnostics Check		OK	01:55

Data Download Date/Time: 02/28/2020 16:03

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001288

INSTRUMENT REGISTERED TO: UNION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDANIEL,MARCUS,G
BREATH TEST OPERATOR AGENCY: Union County SO
ARRESTING OFFICER: HANLON,CARL
ARRESTING OFFICER AGENCY: Union County SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 823252
DRY GAS EXPIRY DATE: 08/07/2020
VIOLATION CODE: DUI
ARREST TIME: 09:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:45

SUBJECT'S LAST NAME: DRUNK

DATE/TIME OF TEST: 02/13/2020 10:28:57

Diagnostics Check		OK	10:17	
Air Blank		0.000	10:18	
Control Test		0.079	10:18	
Air Blank		0.000	10:18	
Subject Sample	# 1		10:21	No Sample Provided
Breath Volume		0.000	10:21	
Air Blank		0.000	10:22	
Air Blank		0.000	10:24	
Subject Sample	# 2		10:27	No Sample Provided
Breath Volume		0.000	10:27	
Air Blank		0.000	10:28	
Control Test		0.080	10:28	
Air Blank		0.000	10:28	
Diagnostics Check		OK	10:28	

Data Download Date/Time: 03/04/2020 12:19

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOWLING,DAKOTA,A
BREATH TEST OPERATOR AGENCY: Cedar Keys PD
ARRESTING OFFICER: DOWLING,DAKOTA
ARRESTING OFFICER AGENCY: Cedar Keys PD
LAST AGENCY INSPECTION: 01/18/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805701
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: PATIN

DATE/TIME OF TEST: 02/02/2020 01:53:38

Diagnostics Check		OK	01:50	
Air Blank		0.000	01:51	
Control Test		0.079	01:51	
Air Blank		0.000	01:52	
Subject Sample	# 1		01:52	Subject Test Refused
Breath Volume		0.000	01:52	
Air Blank		0.000	01:52	
Control Test		0.079	01:53	
Air Blank		0.000	01:53	
Diagnostics Check		OK	01:53	

Data Download Date/Time: 02/15/2020 18:44

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASH,EARL,C
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MASH,EARL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 01/18/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805701
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 12:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:15

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 02/03/2020 13:50:44

Diagnostics Check		OK	13:44
Air Blank		0.000	13:45
Control Test		0.078	13:45
Air Blank		0.000	13:46
Subject Sample	# 1	0.028	13:46
Breath Volume		2.128	13:46
Air Blank		0.000	13:47
Air Blank		0.000	13:48
Subject Sample	# 2	0.026	13:49
Breath Volume		1.476	13:49
Air Blank		0.000	13:49
Control Test		0.078	13:50
Air Blank		0.000	13:50
Diagnostics Check		OK	13:50

Data Download Date/Time: 02/15/2020 18:44

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROE,JEREMIAH,J
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: ROE,JEREMIAH
ARRESTING OFFICER AGENCY: Levy County SO
LAST AGENCY INSPECTION: 02/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG000606
DRY GAS EXPIRY DATE: 01/06/2022
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: MCCALLUM

DATE/TIME OF TEST: 02/26/2020 00:14:34

Diagnostics Check		OK	00:07
Air Blank		0.000	00:07
Control Test		0.079	00:08
Air Blank		0.000	00:08
Subject Sample	# 1	0.070	00:09
Breath Volume		1.761	00:09
Air Blank		0.000	00:10
Air Blank		0.000	00:11
Subject Sample	# 2	0.067	00:13
Breath Volume		2.382	00:13
Air Blank		0.000	00:13
Control Test		0.078	00:13
Air Blank		0.000	00:14
Diagnostics Check		OK	00:14

Data Download Date/Time: 03/28/2020 10:08

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ECKER,DEREK,C
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: REEVES,DAVID
ARRESTING OFFICER AGENCY: Gilchrist County SO
LAST AGENCY INSPECTION: 02/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG000606
DRY GAS EXPIRY DATE: 01/06/2022
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: HILBURN

DATE/TIME OF TEST: 02/28/2020 01:39:22

Diagnostics Check		OK	01:33
Air Blank		0.000	01:33
Control Test		0.080	01:33
Air Blank		0.000	01:34
Subject Sample	# 1	0.159	01:35
Breath Volume		3.542	01:35
Air Blank		0.000	01:35
Air Blank		0.000	01:37
Subject Sample	# 2	0.139	01:37
Breath Volume		1.320	01:37
Air Blank		0.000	01:38
Control Test		0.081	01:38
Air Blank		0.000	01:39
Diagnostics Check		OK	01:39

Data Download Date/Time: 03/28/2020 10:08

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARVOTH,DANIEL,P
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: HARVOTH,DANIEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: SAUNDERS

DATE/TIME OF TEST: 02/29/2020 01:59:59

Diagnostics Check		OK	01:56	
Air Blank		0.000	01:56	
Control Test		0.080	01:56	
Air Blank		0.000	01:57	
Subject Sample	# 1		01:58	Subject Test Refused
Breath Volume		0.000	01:58	
Air Blank		0.000	01:59	
Control Test		0.080	01:59	
Air Blank		0.000	01:59	
Diagnostics Check		OK	01:59	

Data Download Date/Time: 04/02/2020 17:33

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVERA,DANTE,E
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: AVERA,DANTE
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 02/29/2020 03:41:29

Diagnostics Check		OK	03:34
Air Blank		0.000	03:35
Control Test		0.080	03:35
Air Blank		0.000	03:35
Subject Sample	# 1	0.161	03:36
Breath Volume		2.667	03:36
Air Blank		0.000	03:37
Air Blank		0.000	03:39
Subject Sample	# 2	0.151	03:39
Breath Volume		2.226	03:39
Air Blank		0.000	03:40
Control Test		0.078	03:40
Air Blank		0.000	03:41
Diagnostics Check		OK	03:41

Data Download Date/Time: 04/02/2020 17:33

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOLDEN,RYAN,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: AVERA,DANTE
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: BUSBY

DATE/TIME OF TEST: 02/01/2020 02:11:35

Diagnostics Check		OK	02:05
Air Blank		0.000	02:05
Control Test		0.080	02:05
Air Blank		0.000	02:06
Subject Sample	# 1	0.145	02:07
Breath Volume		2.964	02:07
Air Blank		0.000	02:07
Air Blank		0.000	02:09
Subject Sample	# 2	0.144	02:10
Breath Volume		3.597	02:10
Air Blank		0.000	02:10
Control Test		0.079	02:10
Air Blank		0.000	02:11
Diagnostics Check		OK	02:11

Data Download Date/Time: 02/28/2020 19:43

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOLSOM,DAVID,A
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: FOLSOM,DAVID
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: FUGATE

DATE/TIME OF TEST: 02/01/2020 02:38:12

Diagnostics Check	OK	02:37	
Air Blank		02:37	RFI Detect
Air Blank	0.000	02:38	

Data Download Date/Time: 02/28/2020 19:43

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOLSOM,DAVID,A
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: FOLSOM,DAVID
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: FUGATE

DATE/TIME OF TEST: 02/01/2020 02:43:24

Diagnostics Check		OK	02:40	
Air Blank		0.000	02:40	
Control Test		0.080	02:40	
Air Blank		0.000	02:41	
Subject Sample	# 1	0.164	02:41	
Breath Volume		2.093	02:41	
Air Blank			02:42	RFI Detect
Air Blank		0.000	02:42	
Control Test		0.079	02:42	
Air Blank		0.000	02:43	
Diagnostics Check		OK	02:43	

Data Download Date/Time: 02/28/2020 19:43

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER,RAYMOND,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: AVERY,BAYLEE
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 15:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:44

SUBJECT'S LAST NAME: FISH

DATE/TIME OF TEST: 02/01/2020 16:12:29

Diagnostics Check		OK	16:04
Air Blank		0.000	16:05
Control Test		0.080	16:05
Air Blank		0.000	16:06
Subject Sample	# 1	0.246	16:07
Breath Volume		1.769	16:07
Air Blank		0.000	16:08
Air Blank		0.000	16:10
Subject Sample	# 2	0.243	16:10
Breath Volume		2.273	16:10
Air Blank		0.000	16:11
Control Test		0.080	16:11
Air Blank		0.000	16:12
Diagnostics Check		OK	16:12

Data Download Date/Time: 02/28/2020 19:43

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER,RAYMOND,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: LEMMON,TREVOR
ARRESTING OFFICER AGENCY: Milton PD
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 18:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:29

SUBJECT'S LAST NAME: HINOTE

DATE/TIME OF TEST: 02/01/2020 20:00:31

Diagnostics Check		OK	19:56	
Air Blank		0.000	19:56	
Control Test		0.081	19:57	
Air Blank		0.000	19:57	
Subject Sample	# 1		19:59	Subject Test Refused
Breath Volume		0.000	19:59	
Air Blank		0.000	19:59	
Control Test		0.081	19:59	
Air Blank		0.000	20:00	
Diagnostics Check		OK	20:00	

Data Download Date/Time: 02/28/2020 19:43

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMS,KYLE,N
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: SIMMS,KYLE
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 19:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:34

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 02/01/2020 21:01:30

Diagnostics Check		OK	20:55
Air Blank		0.000	20:55
Control Test		0.080	20:55
Air Blank		0.000	20:56
Subject Sample	# 1	0.207	20:57
Breath Volume		2.546	20:57
Air Blank		0.000	20:57
Air Blank		0.000	20:59
Subject Sample	# 2	0.201	20:59
Breath Volume		1.253	20:59
Air Blank		0.000	21:00
Control Test		0.081	21:00
Air Blank		0.000	21:01
Diagnostics Check		OK	21:01

Data Download Date/Time: 02/28/2020 19:43

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCAPIN,BRANDON,L
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: SCAPIN,BRANDON
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 20:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:29

SUBJECT'S LAST NAME: BUCKLEY

DATE/TIME OF TEST: 02/01/2020 21:55:53

Diagnostics Check		OK	21:49
Air Blank		0.000	21:49
Control Test		0.080	21:50
Air Blank		0.000	21:50
Subject Sample	# 1	0.272	21:51
Breath Volume		3.761	21:51
Air Blank		0.000	21:51
Air Blank		0.000	21:53
Subject Sample	# 2	0.264	21:54
Breath Volume		4.257	21:54
Air Blank		0.000	21:54
Control Test		0.080	21:55
Air Blank		0.000	21:55
Diagnostics Check		OK	21:55

Data Download Date/Time: 02/28/2020 19:43

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER,RAYMOND,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SPENCER,RAYMOND
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: KLAUSS

DATE/TIME OF TEST: 02/02/2020 02:13:05

Diagnostics Check		OK	02:06
Air Blank		0.000	02:06
Control Test		0.080	02:06
Air Blank		0.000	02:07
Subject Sample	# 1	0.228	02:08
Breath Volume		3.789	02:08
Air Blank		0.000	02:09
Air Blank		0.000	02:10
Subject Sample	# 2	0.215	02:11
Breath Volume		2.871	02:11
Air Blank		0.000	02:12
Control Test		0.081	02:12
Air Blank		0.000	02:13
Diagnostics Check		OK	02:13

Data Download Date/Time: 02/28/2020 19:43

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,CALEB,Z
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: LYNN,JAMISON
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 06:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:53

SUBJECT'S LAST NAME: ACUELLAR

DATE/TIME OF TEST: 02/02/2020 08:20:26

Diagnostics Check		OK	08:13
Air Blank		0.000	08:14
Control Test		0.080	08:14
Air Blank		0.000	08:15
Subject Sample	# 1	0.223	08:15
Breath Volume		3.570	08:15
Air Blank		0.000	08:16
Air Blank		0.000	08:18
Subject Sample	# 2	0.215	08:18
Breath Volume		3.765	08:18
Air Blank		0.000	08:19
Control Test		0.080	08:19
Air Blank		0.000	08:20
Diagnostics Check		OK	08:20

Data Download Date/Time: 02/28/2020 19:43

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER,RAYMOND,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MOREY,DANIELE
ARRESTING OFFICER AGENCY: Milton PD
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 20:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:07

SUBJECT'S LAST NAME: LONG

DATE/TIME OF TEST: 02/02/2020 22:03:58

Diagnostics Check		OK	21:58	
Air Blank		0.000	21:59	
Control Test		0.081	21:59	
Air Blank		0.000	21:59	
Subject Sample	# 1		22:02	Subject Test Refused
Breath Volume		0.000	22:02	
Air Blank		0.000	22:03	
Control Test		0.081	22:03	
Air Blank		0.000	22:03	
Diagnostics Check		OK	22:03	

Data Download Date/Time: 02/28/2020 19:43

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAPMAN,MAVERICK,E
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CHAPMAN,MAVERICK
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: DEJAYNES

DATE/TIME OF TEST: 02/02/2020 22:25:15

Diagnostics Check		OK	22:18
Air Blank		0.000	22:19
Control Test		0.080	22:19
Air Blank		0.000	22:20
Subject Sample	# 1	0.258	22:20
Breath Volume		3.367	22:20
Air Blank		0.000	22:21
Air Blank		0.000	22:23
Subject Sample	# 2	0.264	22:23
Breath Volume		3.632	22:23
Air Blank		0.000	22:24
Control Test		0.078	22:24
Air Blank		0.000	22:25
Diagnostics Check		OK	22:25

Data Download Date/Time: 02/28/2020 19:43

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAPMAN,MAVERICK,E
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CHAPMAN,MAVERICK
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: GUNTON

DATE/TIME OF TEST: 02/03/2020 00:04:55

Diagnostics Check		OK	23:58
Air Blank		0.000	23:59
Control Test		0.080	23:59
Air Blank		0.000	00:00
Subject Sample	# 1	0.107	00:00
Breath Volume		3.144	00:00
Air Blank		0.000	00:01
Air Blank		0.000	00:02
Subject Sample	# 2	0.101	00:03
Breath Volume		2.867	00:03
Air Blank		0.000	00:04
Control Test		0.080	00:04
Air Blank		0.000	00:04
Diagnostics Check		OK	00:04

Data Download Date/Time: 02/28/2020 19:43

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAPMAN,MAVERICK,E
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: AVERY,BAYLEE
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: BUI

DATE/TIME OF TEST: 02/03/2020 01:15:27

Diagnostics Check		OK	01:11	
Air Blank		0.000	01:11	
Control Test		0.081	01:11	
Air Blank		0.000	01:12	
Subject Sample	# 1		01:14	Subject Test Refused
Breath Volume		0.000	01:14	
Air Blank		0.000	01:14	
Control Test		0.081	01:14	
Air Blank		0.000	01:15	
Diagnostics Check		OK	01:15	

Data Download Date/Time: 02/28/2020 19:43

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARTZOG,R,
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: HENDRIX,T
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 14:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:40

SUBJECT'S LAST NAME: LAY

DATE/TIME OF TEST: 02/04/2020 16:34:43

Diagnostics Check		OK	16:28
Air Blank		0.000	16:28
Control Test		0.080	16:29
Air Blank		0.000	16:29
Subject Sample	# 1	0.184	16:30
Breath Volume		1.964	16:30
Air Blank		0.000	16:30
Air Blank		0.000	16:32
Subject Sample	# 2	0.188	16:33
Breath Volume		3.015	16:33
Air Blank		0.000	16:33
Control Test		0.080	16:34
Air Blank		0.000	16:34
Diagnostics Check		OK	16:34

Data Download Date/Time: 02/28/2020 19:43

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARTZOG,R,
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: POND,CURTIS
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 15:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:15

SUBJECT'S LAST NAME: REEVER

DATE/TIME OF TEST: 02/04/2020 16:47:22

Diagnostics Check		OK	16:39
Air Blank		0.000	16:39
Control Test		0.080	16:39
Air Blank		0.000	16:40
Subject Sample	# 1	0.000	16:40
Breath Volume		1.691	16:40
Air Blank		0.000	16:41
Air Blank		0.000	16:43
Subject Sample	# 2	0.000	16:46
Breath Volume		1.394	16:46
Air Blank		0.000	16:46
Control Test		0.081	16:46
Air Blank		0.000	16:47
Diagnostics Check		OK	16:47

Data Download Date/Time: 02/28/2020 19:43

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURKE,MATTHEW,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: REAVES,CHRISTINA
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 02/06/2020 02:55:11

Diagnostics Check		OK	02:48
Air Blank		0.000	02:48
Control Test		0.080	02:49
Air Blank		0.000	02:49
Subject Sample	# 1	0.000	02:50
Breath Volume		2.078	02:50
Air Blank		0.000	02:51
Air Blank		0.000	02:53
Subject Sample	# 2	0.000	02:53
Breath Volume		1.867	02:53
Air Blank		0.000	02:54
Control Test		0.079	02:54
Air Blank		0.000	02:55
Diagnostics Check		OK	02:55

Data Download Date/Time: 02/28/2020 19:43

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MERRILL,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MERRILL,CHRISTOPER
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 19:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:11

SUBJECT'S LAST NAME: AYRES

DATE/TIME OF TEST: 02/06/2020 20:41:04

Diagnostics Check		OK	20:31	
Air Blank		0.000	20:32	
Control Test		0.080	20:32	
Air Blank		0.000	20:32	
Subject Sample	# 1		20:35	No Sample Provided
Breath Volume		0.000	20:35	
Air Blank		0.000	20:36	
Air Blank		0.000	20:38	
Subject Sample	# 2		20:39	Subject Test Refused
Breath Volume		0.000	20:39	
Air Blank		0.000	20:40	
Control Test		0.080	20:40	
Air Blank		0.000	20:40	
Diagnostics Check		OK	20:41	

Data Download Date/Time: 02/28/2020 19:43

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POSEY,BENJAMIN,E
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: POSEY,BENJAMIN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: POST

DATE/TIME OF TEST: 02/07/2020 01:17:59

Diagnostics Check		OK	01:15	
Air Blank		0.000	01:15	
Control Test		0.080	01:15	
Air Blank		0.000	01:16	
Subject Sample	# 1		01:16	Subject Test Refused
Breath Volume		0.000	01:16	
Air Blank		0.000	01:17	
Control Test		0.080	01:17	
Air Blank		0.000	01:17	
Diagnostics Check		OK	01:17	

Data Download Date/Time: 02/28/2020 19:43

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POSEY,BENJAMIN,E
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: POSEY,BENJAMIN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: POST

DATE/TIME OF TEST: 02/07/2020 01:42:38

Diagnostics Check		OK	01:36
Air Blank		0.000	01:36
Control Test		0.080	01:36
Air Blank		0.000	01:37
Subject Sample	# 1	0.188	01:37
Breath Volume		1.585	01:37
Air Blank		0.000	01:38
Air Blank		0.000	01:40
Subject Sample	# 2	0.195	01:41
Breath Volume		4.738	01:41
Air Blank		0.000	01:41
Control Test		0.079	01:42
Air Blank		0.000	01:42
Diagnostics Check		OK	01:42

Data Download Date/Time: 02/28/2020 19:44

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAPMAN,MAVERICK,E
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CHAPMAN,MAVERICK
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 21:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:28

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 02/08/2020 22:08:22

Diagnostics Check		OK	21:52	
Air Blank		0.000	21:53	
Control Test		0.081	21:53	
Air Blank		0.000	21:54	
Subject Sample	# 1	0.149	21:57	Volume Not Met
Breath Volume		1.085	21:57	
Air Blank		0.000	21:57	
Air Blank		0.000	21:59	
Subject Sample	# 2	0.149	22:01	
Breath Volume		1.378	22:01	
Air Blank		0.000	22:01	
Air Blank		0.000	22:03	
Subject Sample	# 3	0.143	22:06	Slope Not Level
Breath Volume		1.226	22:06	
Air Blank		0.000	22:07	
Control Test		0.079	22:07	
Air Blank		0.000	22:08	
Diagnostics Check		OK	22:08	

Data Download Date/Time: 02/28/2020 19:44

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER,RAYMOND,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SPENCER,RAYMOND
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: ASHCRAFT

DATE/TIME OF TEST: 02/09/2020 03:24:32

Diagnostics Check		OK	03:16
Air Blank		0.000	03:17
Control Test		0.081	03:17
Air Blank		0.000	03:18
Subject Sample	# 1	0.159	03:19
Breath Volume		2.683	03:19
Air Blank		0.000	03:20
Air Blank		0.000	03:22
Subject Sample	# 2	0.154	03:23
Breath Volume		2.476	03:23
Air Blank		0.000	03:23
Control Test		0.080	03:23
Air Blank		0.000	03:24
Diagnostics Check		OK	03:24

Data Download Date/Time: 02/28/2020 19:44

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LARRIVEE-GRANT,CHRISTINA,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: LARRIVEE-GRANT,CHRISTINA
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: MAXWELL

DATE/TIME OF TEST: 02/11/2020 23:01:44

Diagnostics Check		OK	22:56	
Air Blank		0.000	22:56	
Control Test		0.080	22:57	
Air Blank		0.000	22:57	
Subject Sample	# 1		23:00	Subject Test Refused
Breath Volume		0.000	23:00	
Air Blank		0.000	23:00	
Control Test		0.080	23:01	
Air Blank		0.000	23:01	
Diagnostics Check		OK	23:01	

Data Download Date/Time: 02/28/2020 19:44

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREIRA,JOSE,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: PEREIRA,JOSE
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: MCGUFFIN

DATE/TIME OF TEST: 02/12/2020 03:21:26

Diagnostics Check		OK	03:12	
Air Blank		0.000	03:13	
Control Test		0.081	03:13	
Air Blank		0.000	03:14	
Subject Sample	# 1		03:17	No Sample Provided
Breath Volume		0.000	03:17	
Air Blank		0.000	03:17	
Air Blank		0.000	03:19	
Subject Sample	# 2		03:20	Subject Test Refused
Breath Volume		0.000	03:20	
Air Blank		0.000	03:20	
Control Test		0.080	03:20	
Air Blank		0.000	03:21	
Diagnostics Check		OK	03:21	

Data Download Date/Time: 02/28/2020 19:44

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAPMAN,MAVERICK,E
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CHAPMAN,MAVERICK
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: KELLEY

DATE/TIME OF TEST: 02/14/2020 23:15:47

Diagnostics Check		OK	23:09
Air Blank		0.000	23:09
Control Test		0.081	23:10
Air Blank		0.000	23:10
Subject Sample	# 1	0.066	23:11
Breath Volume		1.746	23:11
Air Blank		0.000	23:12
Air Blank		0.000	23:13
Subject Sample	# 2	0.065	23:14
Breath Volume		1.695	23:14
Air Blank		0.000	23:14
Control Test		0.081	23:15
Air Blank		0.000	23:15
Diagnostics Check		OK	23:15

Data Download Date/Time: 02/28/2020 19:44

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,DANIEL,K
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: DAVIS,DANIEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 02/15/2020 00:32:55

Diagnostics Check		OK	00:27
Air Blank		0.000	00:27
Control Test		0.080	00:27
Air Blank		0.000	00:28
Subject Sample	# 1	0.000	00:28
Breath Volume		1.347	00:28
Air Blank		0.000	00:29
Air Blank		0.000	00:31
Subject Sample	# 2	0.000	00:31
Breath Volume		1.371	00:31
Air Blank		0.000	00:32
Control Test		0.081	00:32
Air Blank		0.000	00:32
Diagnostics Check		OK	00:32

Data Download Date/Time: 02/28/2020 19:44

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVIER,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: LEVIER,MATTHEW
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: LIVINGSTON

DATE/TIME OF TEST: 02/15/2020 01:26:45

Diagnostics Check		OK	01:20
Air Blank		0.000	01:21
Control Test		0.082	01:21
Air Blank		0.000	01:21
Subject Sample	# 1	0.203	01:22
Breath Volume		1.746	01:22
Air Blank		0.000	01:22
Air Blank		0.000	01:24
Subject Sample	# 2	0.217	01:25
Breath Volume		2.757	01:25
Air Blank		0.000	01:25
Control Test		0.081	01:26
Air Blank		0.000	01:26
Diagnostics Check		OK	01:26

Data Download Date/Time: 02/28/2020 19:44

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUCE,HEATHER,N
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: BRUCE,HEATHER
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 19:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: CLECKER

DATE/TIME OF TEST: 02/16/2020 19:51:43

Diagnostics Check		OK	19:48	
Air Blank		0.000	19:49	
Control Test		0.081	19:49	
Air Blank		0.000	19:50	
Subject Sample	# 1		19:50	Subject Test Refused
Breath Volume		0.000	19:50	
Air Blank		0.000	19:50	
Control Test		0.081	19:51	
Air Blank		0.000	19:51	
Diagnostics Check		OK	19:51	

Data Download Date/Time: 02/28/2020 19:44

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LARRIVEE-GRANT,CHRISTINA,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: LARRIVEE-GRANT,CHRISTINA
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 18:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: SOBCZAK

DATE/TIME OF TEST: 02/16/2020 20:17:10

Diagnostics Check		OK	20:13	
Air Blank		0.000	20:13	
Control Test		0.081	20:14	
Air Blank		0.000	20:14	
Subject Sample	# 1		20:15	Subject Test Refused
Breath Volume		0.000	20:15	
Air Blank		0.000	20:16	
Control Test		0.081	20:16	
Air Blank		0.000	20:17	
Diagnostics Check		OK	20:17	

Data Download Date/Time: 02/28/2020 19:44

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,ANN,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: GRAY,ANN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 07:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:27

SUBJECT'S LAST NAME: DOBBS

DATE/TIME OF TEST: 02/19/2020 09:08:39

Diagnostics Check		OK	08:57	
Air Blank		0.000	08:57	
Control Test		0.080	08:58	
Air Blank		0.000	08:58	
Subject Sample	# 1		09:01	No Sample Provided
Breath Volume		0.000	09:01	
Air Blank		0.000	09:02	
Air Blank		0.000	09:04	
Subject Sample	# 2		09:07	No Sample Provided
Breath Volume		0.000	09:07	
Air Blank		0.000	09:07	
Control Test		0.081	09:08	
Air Blank		0.000	09:08	
Diagnostics Check		OK	09:08	

Data Download Date/Time: 02/28/2020 19:44

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOBLES,DARIUS,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: NOBLES,DARIUS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 20:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: LONGMORE

DATE/TIME OF TEST: 02/19/2020 21:31:16

Diagnostics Check		OK	21:24
Air Blank		0.000	21:24
Control Test		0.081	21:25
Air Blank		0.000	21:25
Subject Sample	# 1	0.191	21:26
Breath Volume		3.320	21:26
Air Blank		0.000	21:27
Air Blank		0.000	21:29
Subject Sample	# 2	0.180	21:29
Breath Volume		2.218	21:29
Air Blank		0.000	21:30
Control Test		0.080	21:30
Air Blank		0.000	21:31
Diagnostics Check		OK	21:31

Data Download Date/Time: 02/28/2020 19:44

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,COREY,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: JONES,COREY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 21:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: ULLRICH

DATE/TIME OF TEST: 02/20/2020 22:43:14

Diagnostics Check		OK	22:36
Air Blank		0.000	22:37
Control Test		0.081	22:37
Air Blank		0.000	22:37
Subject Sample	# 1	0.118	22:38
Breath Volume		3.730	22:38
Air Blank		0.000	22:39
Air Blank		0.000	22:41
Subject Sample	# 2	0.113	22:41
Breath Volume		2.816	22:41
Air Blank		0.000	22:42
Control Test		0.079	22:42
Air Blank		0.000	22:43
Diagnostics Check		OK	22:43

Data Download Date/Time: 02/28/2020 19:44

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAPMAN,MAVERICK,E
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CHAPMAN,MAVERICK
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 16:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:55

SUBJECT'S LAST NAME: CREEL

DATE/TIME OF TEST: 02/21/2020 17:22:53

Diagnostics Check		OK	17:18	
Air Blank		0.000	17:18	
Control Test		0.081	17:18	
Air Blank		0.000	17:19	
Subject Sample	# 1		17:21	Subject Test Refused
Breath Volume		0.000	17:21	
Air Blank		0.000	17:21	
Control Test		0.082	17:22	
Air Blank		0.000	17:22	
Diagnostics Check		OK	17:22	

Data Download Date/Time: 02/28/2020 19:44

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARLSON,JOSHUA,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: AVERY,BAYLEE
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 19:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:23

SUBJECT'S LAST NAME: NIX

DATE/TIME OF TEST: 02/21/2020 20:51:47

Diagnostics Check		OK	20:45
Air Blank		0.000	20:45
Control Test		0.082	20:46
Air Blank		0.000	20:46
Subject Sample	# 1	0.000	20:47
Breath Volume		2.566	20:47
Air Blank		0.000	20:47
Air Blank		0.000	20:49
Subject Sample	# 2	0.000	20:50
Breath Volume		2.421	20:50
Air Blank		0.000	20:50
Control Test		0.082	20:51
Air Blank		0.000	20:51
Diagnostics Check		OK	20:51

Data Download Date/Time: 02/28/2020 19:44

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWELL,DAVID,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ROWELL,DAVID
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: PITTS

DATE/TIME OF TEST: 02/21/2020 21:27:42

Diagnostics Check		OK	21:20
Air Blank		0.000	21:20
Control Test		0.082	21:20
Air Blank		0.000	21:21
Subject Sample	# 1	0.204	21:23
Breath Volume		3.167	21:23
Air Blank		0.000	21:23
Air Blank		0.000	21:25
Subject Sample	# 2	0.196	21:26
Breath Volume		2.882	21:26
Air Blank		0.000	21:26
Control Test		0.080	21:27
Air Blank		0.000	21:27
Diagnostics Check		OK	21:27

Data Download Date/Time: 02/28/2020 19:44

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVIER,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: DAVIS,NATHANIEL
ARRESTING OFFICER AGENCY: Milton PD
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: KELLEY

DATE/TIME OF TEST: 02/22/2020 03:21:06

Diagnostics Check		OK	03:14
Air Blank		0.000	03:15
Control Test		0.081	03:15
Air Blank		0.000	03:16
Subject Sample	# 1	0.148	03:16
Breath Volume		3.316	03:16
Air Blank		0.000	03:17
Air Blank		0.000	03:19
Subject Sample	# 2	0.148	03:19
Breath Volume		3.070	03:19
Air Blank		0.000	03:20
Control Test		0.081	03:20
Air Blank		0.000	03:21
Diagnostics Check		OK	03:21

Data Download Date/Time: 02/28/2020 19:44

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNSFORD,WILLIAM,C
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: DUNSFORD,WILLIAM
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: CONTANT

DATE/TIME OF TEST: 02/22/2020 04:22:45

Diagnostics Check		OK	04:16
Air Blank		0.000	04:16
Control Test		0.080	04:16
Air Blank		0.000	04:17
Subject Sample	# 1	0.122	04:18
Breath Volume		2.238	04:18
Air Blank		0.000	04:18
Air Blank		0.000	04:20
Subject Sample	# 2	0.127	04:21
Breath Volume		2.343	04:21
Air Blank		0.000	04:21
Control Test		0.079	04:22
Air Blank		0.000	04:22
Diagnostics Check		OK	04:22

Data Download Date/Time: 02/28/2020 19:44

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARTZOG,ROBERT,
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: POND,CURTIS
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 17:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: MCGRAW

DATE/TIME OF TEST: 02/23/2020 18:45:56

Diagnostics Check		OK	18:34	
Air Blank		0.000	18:34	
Control Test		0.082	18:35	
Air Blank		0.000	18:35	
Subject Sample	# 1	0.096	18:38	Volume Not Met
Breath Volume		0.031	18:38	
Air Blank		0.000	18:39	
Air Blank		0.000	18:41	
Subject Sample	# 2	0.113	18:44	Volume Not Met
Breath Volume		0.027	18:44	
Air Blank		0.000	18:45	
Control Test		0.080	18:45	
Air Blank		0.000	18:45	
Diagnostics Check		OK	18:45	

Data Download Date/Time: 02/28/2020 19:44

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCAPIN,BRANDON,L
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: SCAPIN,BRANDON
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: SETTLES

DATE/TIME OF TEST: 02/27/2020 03:54:32

Diagnostics Check		OK	03:47
Air Blank		0.000	03:48
Control Test		0.082	03:48
Air Blank		0.000	03:49
Subject Sample	# 1	0.168	03:49
Breath Volume		3.328	03:49
Air Blank		0.000	03:50
Air Blank		0.000	03:52
Subject Sample	# 2	0.173	03:53
Breath Volume		3.300	03:53
Air Blank		0.000	03:53
Control Test		0.081	03:53
Air Blank		0.000	03:54
Diagnostics Check		OK	03:54

Data Download Date/Time: 02/28/2020 19:44

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: WEBER,JEFFREY
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:04

SUBJECT'S LAST NAME: APLIN

DATE/TIME OF TEST: 02/01/2020 05:32:55

Diagnostics Check		OK	05:26
Air Blank		0.000	05:27
Control Test		0.080	05:27
Air Blank		0.000	05:28
Subject Sample	# 1	0.206	05:28
Breath Volume		1.847	05:28
Air Blank		0.000	05:29
Air Blank		0.000	05:31
Subject Sample	# 2	0.199	05:31
Breath Volume		1.425	05:31
Air Blank		0.000	05:32
Control Test		0.081	05:32
Air Blank		0.000	05:32
Diagnostics Check		OK	05:32

Data Download Date/Time: 02/27/2020 10:36

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUTTON,KERRY,D
BREATH TEST OPERATOR AGENCY: Lynn Haven PD
ARRESTING OFFICER: DUTTON,KERRY
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: MINOR

DATE/TIME OF TEST: 02/09/2020 01:38:42

Diagnostics Check		OK	01:35	
Air Blank		0.000	01:36	
Control Test		0.081	01:36	
Air Blank		0.000	01:37	
Subject Sample	# 1		01:37	Subject Test Refused
Breath Volume		0.000	01:37	
Air Blank		0.000	01:37	
Control Test		0.081	01:38	
Air Blank		0.000	01:38	
Diagnostics Check		OK	01:38	

Data Download Date/Time: 02/27/2020 10:36

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRELL,CHAD,A
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: BRANCH,JOSH
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 21:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 02/18/2020 22:22:26

Diagnostics Check		OK	22:16
Air Blank		0.000	22:16
Control Test		0.081	22:17
Air Blank		0.000	22:17
Subject Sample	# 1	0.182	22:18
Breath Volume		1.906	22:18
Air Blank		0.000	22:18
Air Blank		0.000	22:20
Subject Sample	# 2	0.181	22:20
Breath Volume		1.992	22:20
Air Blank		0.000	22:21
Control Test		0.079	22:21
Air Blank		0.000	22:22
Diagnostics Check		OK	22:22

Data Download Date/Time: 02/27/2020 10:36

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FURMAN,FREDDIE,R
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: PRICE,CARLON
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 19:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:54

SUBJECT'S LAST NAME: HOOVER

DATE/TIME OF TEST: 02/19/2020 21:20:27

Diagnostics Check		OK	21:14
Air Blank		0.000	21:14
Control Test		0.081	21:15
Air Blank		0.000	21:15
Subject Sample	# 1	0.000	21:16
Breath Volume		2.402	21:16
Air Blank		0.000	21:16
Air Blank		0.000	21:18
Subject Sample	# 2	0.000	21:19
Breath Volume		2.527	21:19
Air Blank		0.000	21:19
Control Test		0.081	21:19
Air Blank		0.000	21:20
Diagnostics Check		OK	21:20

Data Download Date/Time: 02/27/2020 10:36

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRELL,CHAD,A
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: ADAMS,HOLTON
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 18:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: MEADOWS

DATE/TIME OF TEST: 02/21/2020 19:46:57

Diagnostics Check		OK	19:40
Air Blank		0.000	19:40
Control Test		0.080	19:41
Air Blank		0.000	19:41
Subject Sample	# 1	0.229	19:42
Breath Volume		2.292	19:42
Air Blank		0.000	19:42
Air Blank		0.000	19:44
Subject Sample	# 2	0.221	19:45
Breath Volume		1.898	19:45
Air Blank		0.000	19:46
Control Test		0.081	19:46
Air Blank		0.000	19:46
Diagnostics Check		OK	19:46

Data Download Date/Time: 02/27/2020 10:36

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWMAN,JOANNA,N
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MCNEIL,RAYMOND
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 02/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG923908
DRY GAS EXPIRY DATE: 08/27/2021
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 02/28/2020 01:17:55

Diagnostics Check		OK	01:11
Air Blank		0.000	01:12
Control Test		0.081	01:12
Air Blank		0.000	01:13
Subject Sample	# 1	0.173	01:13
Breath Volume		1.375	01:13
Air Blank		0.000	01:14
Air Blank		0.000	01:16
Subject Sample	# 2	0.167	01:16
Breath Volume		1.468	01:16
Air Blank		0.000	01:17
Control Test		0.081	01:17
Air Blank		0.000	01:17
Diagnostics Check		OK	01:17

Data Download Date/Time: 03/24/2020 15:39

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARRICO,KELLY,D
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: CARRICO,KELLY
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 01/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 869968
DRY GAS EXPIRY DATE: 12/06/2020
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: GMYTRUK

DATE/TIME OF TEST: 02/02/2020 23:46:28

Diagnostics Check		OK	23:38
Air Blank		0.000	23:38
Control Test		0.078	23:38
Air Blank		0.000	23:39
Subject Sample	# 1	0.169	23:41
Breath Volume		1.824	23:41
Air Blank		0.000	23:41
Air Blank		0.000	23:43
Subject Sample	# 2	0.162	23:44
Breath Volume		1.589	23:44
Air Blank		0.000	23:45
Control Test		0.078	23:45
Air Blank		0.000	23:46
Diagnostics Check		OK	23:46

Data Download Date/Time: 02/21/2020 17:38

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POVEDA,CARLOS,M
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: POVEDA,CARLOS
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 01/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 869968
DRY GAS EXPIRY DATE: 12/06/2020
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: COURTNEY

DATE/TIME OF TEST: 02/16/2020 02:43:42

Diagnostics Check		OK	02:41	
Air Blank		0.000	02:41	
Control Test		0.078	02:42	
Air Blank		0.000	02:42	
Subject Sample	# 1		02:43	Subject Test Refused
Breath Volume		0.000	02:43	
Air Blank			02:43	RFI Detect
Air Blank		0.000	02:43	

Data Download Date/Time: 02/21/2020 17:38

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPRAGUE,CHARLES,T
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: SPRAGUE,CHARLES
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 01/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 869968
DRY GAS EXPIRY DATE: 12/06/2020
VIOLATION CODE: DUI
ARREST TIME: 22:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: KELLER

DATE/TIME OF TEST: 02/20/2020 23:13:52

Diagnostics Check		OK	23:06
Air Blank		0.000	23:06
Control Test		0.077	23:07
Air Blank		0.000	23:07
Subject Sample	# 1	0.000	23:09
Breath Volume		2.230	23:09
Air Blank		0.000	23:09
Air Blank		0.000	23:11
Subject Sample	# 2	0.000	23:12
Breath Volume		1.210	23:12
Air Blank		0.000	23:12
Control Test		0.076	23:13
Air Blank		0.000	23:13
Diagnostics Check		OK	23:13

Data Download Date/Time: 02/21/2020 17:38

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSE,ALEXIS,M
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: SCALIA,A
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 02/01/2020 22:39:14

Diagnostics Check		OK	22:33
Air Blank		0.000	22:33
Control Test		0.080	22:33
Air Blank		0.000	22:34
Subject Sample	# 1	0.187	22:34
Breath Volume		1.589	22:34
Air Blank		0.000	22:35
Air Blank		0.000	22:37
Subject Sample	# 2	0.174	22:37
Breath Volume		1.222	22:37
Air Blank		0.000	22:38
Control Test		0.079	22:38
Air Blank		0.000	22:39
Diagnostics Check		OK	22:39

Data Download Date/Time: 02/27/2020 06:47

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYOUNGHANSE,KYLE,O
BREATH TEST OPERATOR AGENCY: DeFuniak Springs PD
ARRESTING OFFICER: ROCKETT,JOHNATHAN
ARRESTING OFFICER AGENCY: DeFuniak Springs PD
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 02/14/2020 03:00:30

Diagnostics Check		OK	02:54
Air Blank		0.000	02:54
Control Test		0.082	02:54
Air Blank		0.000	02:55
Subject Sample	# 1	0.227	02:55
Breath Volume		3.394	02:55
Air Blank		0.000	02:56
Air Blank		0.000	02:58
Subject Sample	# 2	0.228	02:58
Breath Volume		3.390	02:58
Air Blank		0.000	02:59
Control Test		0.080	02:59
Air Blank		0.000	03:00
Diagnostics Check		OK	03:00

Data Download Date/Time: 02/27/2020 06:47

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,ANDREW,S
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: CAREY,ANDREW
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:25

SUBJECT'S LAST NAME: BUENTELLO MARTINEZ

DATE/TIME OF TEST: 02/16/2020 06:02:32

Diagnostics Check		OK	05:55
Air Blank		0.000	05:56
Control Test		0.080	05:56
Air Blank		0.000	05:57
Subject Sample	# 1	0.157	05:57
Breath Volume		1.753	05:57
Air Blank		0.000	05:58
Air Blank		0.000	06:00
Subject Sample	# 2	0.159	06:00
Breath Volume		1.914	06:00
Air Blank		0.000	06:01
Control Test		0.081	06:01
Air Blank		0.000	06:02
Diagnostics Check		OK	06:02

Data Download Date/Time: 02/27/2020 06:48

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENNETT,DAVID,J
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: MAY,WILLIAM
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 02/17/2020 02:16:27

Diagnostics Check		OK	02:10
Air Blank		0.000	02:10
Control Test		0.079	02:11
Air Blank		0.000	02:11
Subject Sample	# 1	0.176	02:12
Breath Volume		1.941	02:12
Air Blank		0.000	02:12
Air Blank		0.000	02:14
Subject Sample	# 2	0.159	02:14
Breath Volume		1.621	02:14
Air Blank		0.000	02:15
Control Test		0.079	02:15
Air Blank		0.000	02:16
Diagnostics Check		OK	02:16

Data Download Date/Time: 02/27/2020 06:48

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYOUNGHANSE,KMYLE,O
BREATH TEST OPERATOR AGENCY: DeFuniak Springs PD
ARRESTING OFFICER: MYOUNGHANSE,KMYLE
ARRESTING OFFICER AGENCY: DeFuniak Springs PD
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: PEVTERS

DATE/TIME OF TEST: 02/17/2020 23:42:49

Diagnostics Check		OK	23:36
Air Blank		0.000	23:36
Control Test		0.079	23:37
Air Blank		0.000	23:37
Subject Sample	# 1	0.172	23:38
Breath Volume		3.062	23:38
Air Blank		0.000	23:38
Air Blank		0.000	23:40
Subject Sample	# 2	0.158	23:41
Breath Volume		2.910	23:41
Air Blank		0.000	23:41
Control Test		0.078	23:42
Air Blank		0.000	23:42
Diagnostics Check		OK	23:42

Data Download Date/Time: 02/27/2020 06:48

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,MARK,A
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: VULETA,B
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 01/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: HARVEY

DATE/TIME OF TEST: 02/01/2020 03:27:05

Diagnostics Check		OK	03:24	
Air Blank		0.000	03:24	
Control Test		0.081	03:24	
Air Blank		0.000	03:25	
Subject Sample	# 1		03:25	Subject Test Refused
Breath Volume		0.000	03:25	
Air Blank		0.000	03:26	
Control Test		0.081	03:26	
Air Blank		0.000	03:27	
Diagnostics Check		OK	03:27	

Data Download Date/Time: 02/06/2020 17:38

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,MARK,A
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: VULETA,B
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 01/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 02/02/2020 00:06:22

Diagnostics Check		OK	00:03	
Air Blank		0.000	00:03	
Control Test		0.081	00:03	
Air Blank		0.000	00:04	
Subject Sample	# 1		00:05	Subject Test Refused
Breath Volume		0.000	00:05	
Air Blank		0.000	00:05	
Control Test		0.081	00:05	
Air Blank		0.000	00:06	
Diagnostics Check		OK	00:06	

Data Download Date/Time: 02/06/2020 17:39

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,STEVEN,D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK,STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 02/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: BUDOVSKY

DATE/TIME OF TEST: 02/05/2020 23:11:35

Diagnostics Check		OK	23:08	
Air Blank		0.000	23:08	
Control Test		0.080	23:09	
Air Blank		0.000	23:09	
Subject Sample	# 1		23:10	Subject Test Refused
Breath Volume		0.000	23:10	
Air Blank		0.000	23:10	
Control Test		0.080	23:10	
Air Blank		0.000	23:11	
Diagnostics Check		OK	23:11	

Data Download Date/Time: 02/06/2020 17:39

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,STEVEN,D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK,STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 02/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 16:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:16

SUBJECT'S LAST NAME: VANHOUTEN

DATE/TIME OF TEST: 02/07/2020 17:42:29

Diagnostics Check		OK	17:36
Air Blank		0.000	17:36
Control Test		0.081	17:37
Air Blank		0.000	17:37
Subject Sample	# 1	0.000	17:38
Breath Volume		1.808	17:38
Air Blank		0.000	17:38
Air Blank		0.000	17:40
Subject Sample	# 2	0.000	17:41
Breath Volume		1.972	17:41
Air Blank		0.000	17:41
Control Test		0.081	17:41
Air Blank		0.000	17:42
Diagnostics Check		OK	17:42

Data Download Date/Time: 03/05/2020 09:34

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,MARK,A
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: VULETA,B
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 02/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 13:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:51

SUBJECT'S LAST NAME: COCKFIELD

DATE/TIME OF TEST: 02/11/2020 14:17:17

Diagnostics Check		OK	14:11
Air Blank		0.000	14:11
Control Test		0.080	14:12
Air Blank		0.000	14:12
Subject Sample	# 1	0.000	14:13
Breath Volume		2.511	14:13
Air Blank		0.000	14:13
Air Blank		0.000	14:15
Subject Sample	# 2	0.000	14:15
Breath Volume		2.097	14:15
Air Blank		0.000	14:16
Control Test		0.080	14:16
Air Blank		0.000	14:17
Diagnostics Check		OK	14:17

Data Download Date/Time: 03/05/2020 09:34

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,CARL,C
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: VARNADORE,RICHARD
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 02/04/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 14:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:59

SUBJECT'S LAST NAME: CINDRIC

DATE/TIME OF TEST: 02/13/2020 15:26:05

Diagnostics Check		OK	15:19
Air Blank		0.000	15:20
Control Test		0.080	15:20
Air Blank		0.000	15:21
Subject Sample	# 1	0.184	15:21
Breath Volume		1.914	15:21
Air Blank		0.000	15:22
Air Blank		0.000	15:24
Subject Sample	# 2	0.176	15:24
Breath Volume		1.273	15:24
Air Blank		0.000	15:25
Control Test		0.079	15:25
Air Blank		0.000	15:26
Diagnostics Check		OK	15:26

Data Download Date/Time: 03/05/2020 09:34

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,STEVEN,D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK,STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 02/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 20:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:02

SUBJECT'S LAST NAME: MEDINA VIRAMONTES

DATE/TIME OF TEST: 02/13/2020 21:28:52

Diagnostics Check		OK	21:22
Air Blank		0.000	21:23
Control Test		0.080	21:23
Air Blank		0.000	21:23
Subject Sample	# 1	0.203	21:24
Breath Volume		2.652	21:24
Air Blank		0.000	21:25
Air Blank		0.000	21:26
Subject Sample	# 2	0.197	21:27
Breath Volume		1.890	21:27
Air Blank		0.000	21:27
Control Test		0.080	21:28
Air Blank		0.000	21:28
Diagnostics Check		OK	21:28

Data Download Date/Time: 03/05/2020 09:34

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,STEVEN,D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK,STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 02/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 19:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:44

SUBJECT'S LAST NAME: JANSEN

DATE/TIME OF TEST: 02/14/2020 20:07:11

Diagnostics Check		OK	20:04	
Air Blank		0.000	20:04	
Control Test		0.081	20:05	
Air Blank		0.000	20:05	
Subject Sample	# 1		20:05	Subject Test Refused
Breath Volume		0.000	20:05	
Air Blank		0.000	20:06	
Control Test		0.081	20:06	
Air Blank		0.000	20:07	
Diagnostics Check		OK	20:07	

Data Download Date/Time: 03/05/2020 09:34

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,STEVEN,D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK,STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 02/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 19:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:26

SUBJECT'S LAST NAME: JIMENEZ

DATE/TIME OF TEST: 02/15/2020 20:01:07

Diagnostics Check		OK	19:46	
Air Blank		0.000	19:47	
Control Test		0.081	19:47	
Air Blank		0.000	19:47	
Subject Sample	# 1	0.216	19:51	Volume Not Met
Breath Volume		0.140	19:51	
Air Blank		0.000	19:51	
Air Blank		0.000	19:53	
Subject Sample	# 2	0.241	19:56	No .020 agreement
Breath Volume		1.718	19:56	
Air Blank		0.000	19:56	
Air Blank		0.000	19:58	
Subject Sample	# 3	0.268	19:59	No .020 agreement
Breath Volume		2.847	19:59	
Air Blank		0.000	20:00	
Control Test		0.078	20:00	
Air Blank		0.000	20:01	
Diagnostics Check		OK	20:01	

Data Download Date/Time: 03/05/2020 09:34

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,STEVEN,D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK,STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 02/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: CHAVEZ MEDINA

DATE/TIME OF TEST: 02/22/2020 01:52:59

Diagnostics Check		OK	01:46
Air Blank		0.000	01:46
Control Test		0.081	01:46
Air Blank		0.000	01:47
Subject Sample	# 1	0.110	01:48
Breath Volume		1.597	01:48
Air Blank		0.000	01:49
Air Blank		0.000	01:51
Subject Sample	# 2	0.100	01:51
Breath Volume		1.777	01:51
Air Blank		0.000	01:52
Control Test		0.081	01:52
Air Blank		0.000	01:52
Diagnostics Check		OK	01:52

Data Download Date/Time: 03/05/2020 09:34

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,STEVEN,D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK,STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 02/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: D\AGUILAR

DATE/TIME OF TEST: 02/23/2020 02:07:52

Diagnostics Check		OK	02:01
Air Blank		0.000	02:02
Control Test		0.082	02:02
Air Blank		0.000	02:02
Subject Sample	# 1	0.134	02:03
Breath Volume		2.222	02:03
Air Blank		0.000	02:03
Air Blank		0.000	02:05
Subject Sample	# 2	0.132	02:06
Breath Volume		2.480	02:06
Air Blank		0.000	02:06
Control Test		0.081	02:07
Air Blank		0.000	02:07
Diagnostics Check		OK	02:07

Data Download Date/Time: 03/05/2020 09:34

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCINERNEY,JESSE,L
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: MCINERNEY,JESS
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 02/08/2020 03:30:15

Diagnostics Check		OK	03:22
Air Blank		0.000	03:22
Control Test		0.080	03:23
Air Blank		0.000	03:23
Subject Sample	# 1	0.179	03:25
Breath Volume		1.367	03:25
Air Blank		0.000	03:26
Air Blank		0.000	03:27
Subject Sample	# 2	0.163	03:28
Breath Volume		2.109	03:28
Air Blank		0.000	03:29
Control Test		0.078	03:29
Air Blank		0.000	03:30
Diagnostics Check		OK	03:30

Data Download Date/Time: 02/27/2020 16:10

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCINERNEY,JESSE,L
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: REISINGER,COLTON
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: DOLNICK

DATE/TIME OF TEST: 02/11/2020 02:32:24

Diagnostics Check		OK	02:25
Air Blank		0.000	02:26
Control Test		0.079	02:26
Air Blank		0.000	02:27
Subject Sample	# 1	0.163	02:27
Breath Volume		2.085	02:27
Air Blank		0.000	02:28
Air Blank		0.000	02:30
Subject Sample	# 2	0.176	02:30
Breath Volume		2.328	02:30
Air Blank		0.000	02:31
Control Test		0.079	02:31
Air Blank		0.000	02:32
Diagnostics Check		OK	02:32

Data Download Date/Time: 02/27/2020 16:10

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TREVISOL,JOSEPH,D
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: HERZOG,DEAN
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:51

SUBJECT'S LAST NAME: VILLA

DATE/TIME OF TEST: 02/14/2020 19:17:59

Diagnostics Check		OK	19:11
Air Blank		0.000	19:12
Control Test		0.079	19:12
Air Blank		0.000	19:12
Subject Sample	# 1	0.203	19:13
Breath Volume		2.648	19:13
Air Blank		0.000	19:14
Air Blank		0.000	19:15
Subject Sample	# 2	0.188	19:16
Breath Volume		2.812	19:16
Air Blank		0.000	19:17
Control Test		0.078	19:17
Air Blank		0.000	19:17
Diagnostics Check		OK	19:17

Data Download Date/Time: 02/27/2020 16:11

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCINERNEY,JESSE,L
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: MCINERNEY,JESSE
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: ECHEANDIA

DATE/TIME OF TEST: 02/17/2020 04:26:36

Diagnostics Check		OK	04:18
Air Blank		0.000	04:19
Control Test		0.079	04:19
Air Blank		0.000	04:20
Subject Sample	# 1	0.180	04:21
Breath Volume		4.289	04:21
Air Blank		0.000	04:22
Air Blank		0.000	04:24
Subject Sample	# 2	0.163	04:25
Breath Volume		2.816	04:25
Air Blank		0.000	04:25
Control Test		0.077	04:26
Air Blank		0.000	04:26
Diagnostics Check		OK	04:26

Data Download Date/Time: 02/27/2020 16:11

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCINERNEY,JESSE,L
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: MCINERNEY,JESSE
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: FAIR

DATE/TIME OF TEST: 02/19/2020 03:20:53

Diagnostics Check		OK	03:12
Air Blank		0.000	03:13
Control Test		0.080	03:13
Air Blank		0.000	03:14
Subject Sample	# 1	0.211	03:15
Breath Volume		2.695	03:15
Air Blank		0.000	03:16
Air Blank		0.000	03:18
Subject Sample	# 2	0.206	03:19
Breath Volume		2.085	03:19
Air Blank		0.000	03:19
Control Test		0.078	03:20
Air Blank		0.000	03:20
Diagnostics Check		OK	03:20

Data Download Date/Time: 02/27/2020 16:11

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,MICHAEL,C
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: GASPAR,BENJAMIN
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 02/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: FUNES

DATE/TIME OF TEST: 02/27/2020 23:35:45

Diagnostics Check		OK	23:28	
Air Blank		0.000	23:28	
Control Test		0.080	23:29	
Air Blank		0.000	23:29	
Subject Sample	# 1		23:32	No Sample Provided
Breath Volume		0.000	23:32	
Air Blank		0.000	23:33	
Air Blank			23:35	RFI Detect
Air Blank			23:35	RFI Detect
Air Blank		0.000	23:35	

Data Download Date/Time: 03/31/2020 16:02

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KANE,JAMES,G
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: BUFFINGTON,BEN
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 14:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:30

SUBJECT'S LAST NAME: SANTANIELLO

DATE/TIME OF TEST: 02/19/2020 15:03:37

Diagnostics Check		OK	15:00	
Air Blank		0.000	15:00	
Control Test		0.080	15:01	
Air Blank		0.000	15:01	
Subject Sample	# 1		15:02	Interferent Detect
Breath Volume		2.535	15:02	
Air Blank		0.000	15:02	
Control Test		0.080	15:03	
Air Blank		0.000	15:03	
Diagnostics Check		OK	15:03	

Data Download Date/Time: 03/06/2020 04:13

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KANE,JAMES,G
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: BUFFINGTON,BEN
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 14:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:30

SUBJECT'S LAST NAME: SANTANIELLO

DATE/TIME OF TEST: 02/19/2020 15:09:16

Diagnostics Check		OK	15:06	
Air Blank		0.000	15:06	
Control Test		0.080	15:06	
Air Blank		0.000	15:07	
Subject Sample	# 1		15:07	Interferent Detect
Breath Volume		2.593	15:07	
Air Blank		0.000	15:08	
Control Test		0.080	15:08	
Air Blank		0.000	15:09	
Diagnostics Check		OK	15:09	

Data Download Date/Time: 03/06/2020 04:13

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HISS,J,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: DAUGHERTY,P
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: SEGMAN

DATE/TIME OF TEST: 02/01/2020 22:20:26

Diagnostics Check		OK	22:12
Air Blank		0.000	22:13
Control Test		0.083	22:13
Air Blank		0.000	22:14
Subject Sample	# 1	0.260	22:15
Breath Volume		1.253	22:15
Air Blank		0.000	22:16
Air Blank		0.000	22:17
Subject Sample	# 2	0.257	22:18
Breath Volume		1.117	22:18
Air Blank		0.000	22:19
Control Test		0.081	22:19
Air Blank		0.000	22:20
Diagnostics Check		OK	22:20

Data Download Date/Time: 02/07/2020 08:36

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIGNORELLI,A,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: GAST,E
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 18:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:14

SUBJECT'S LAST NAME: WISE

DATE/TIME OF TEST: 02/03/2020 19:46:50

Diagnostics Check		OK	19:39
Air Blank		0.000	19:39
Control Test		0.084	19:40
Air Blank		0.000	19:40
Subject Sample	# 1	0.181	19:41
Breath Volume		2.960	19:41
Air Blank		0.000	19:42
Air Blank		0.000	19:44
Subject Sample	# 2	0.175	19:45
Breath Volume		2.593	19:45
Air Blank		0.000	19:45
Control Test		0.082	19:46
Air Blank		0.000	19:46
Diagnostics Check		OK	19:46

Data Download Date/Time: 02/07/2020 08:36

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIGNORELI,A,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: INGRAM,J
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 01/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 20:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:49

SUBJECT'S LAST NAME: CHIRCO

DATE/TIME OF TEST: 02/03/2020 21:25:06

Diagnostics Check		OK	21:22	
Air Blank		0.000	21:22	
Control Test		0.084	21:23	
Air Blank		0.000	21:23	
Subject Sample	# 1		21:23	Subject Test Refused
Breath Volume		0.000	21:23	
Air Blank		0.000	21:24	
Control Test		0.083	21:24	
Air Blank		0.000	21:25	
Diagnostics Check		OK	21:25	

Data Download Date/Time: 02/07/2020 08:36

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIGNORELLI,A,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: OREHOSTKY,G
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 19:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: MCCARTNEY

DATE/TIME OF TEST: 02/03/2020 21:58:49

Diagnostics Check		OK	21:52
Air Blank		0.000	21:52
Control Test		0.083	21:52
Air Blank		0.000	21:53
Subject Sample	# 1	0.000	21:54
Breath Volume		2.023	21:54
Air Blank		0.000	21:54
Air Blank		0.000	21:56
Subject Sample	# 2	0.000	21:57
Breath Volume		2.234	21:57
Air Blank		0.000	21:57
Control Test		0.083	21:58
Air Blank		0.000	21:58
Diagnostics Check		OK	21:58

Data Download Date/Time: 02/07/2020 08:36

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RICKY,
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: PEREZ,RICKY
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 01/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: CORNELL

DATE/TIME OF TEST: 02/04/2020 01:24:47

Diagnostics Check		OK	01:10	
Air Blank		0.000	01:11	
Control Test		0.082	01:11	
Air Blank		0.000	01:11	
Subject Sample	# 1	0.085	01:15	Volume Not Met
Breath Volume		0.316	01:15	
Air Blank		0.000	01:15	
Air Blank		0.000	01:17	
Subject Sample	# 2	0.105	01:19	
Breath Volume		1.238	01:19	
Air Blank		0.000	01:20	
Air Blank		0.000	01:21	
Subject Sample	# 3	0.098	01:23	
Breath Volume		1.375	01:23	
Air Blank		0.000	01:23	
Control Test		0.082	01:24	
Air Blank		0.000	01:24	
Diagnostics Check		OK	01:24	

Data Download Date/Time: 02/07/2020 08:36

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAYMOND,FRANK,W
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: VILLIER,D/S
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 14:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:45

SUBJECT'S LAST NAME: DAVISON

DATE/TIME OF TEST: 02/05/2020 15:23:39

Diagnostics Check		OK	15:19	
Air Blank		0.000	15:20	
Control Test		0.084	15:20	
Air Blank		0.000	15:20	
Subject Sample	# 1		15:22	Subject Test Refused
Breath Volume		0.000	15:22	
Air Blank		0.000	15:22	
Control Test		0.083	15:23	
Air Blank		0.000	15:23	
Diagnostics Check		OK	15:23	

Data Download Date/Time: 02/07/2020 08:36

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHILE,C,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: RUIZ,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:58

SUBJECT'S LAST NAME: FASOLI

DATE/TIME OF TEST: 02/05/2020 21:28:08

Diagnostics Check		OK	21:21
Air Blank		0.000	21:21
Control Test		0.082	21:22
Air Blank		0.000	21:22
Subject Sample	# 1	0.073	21:23
Breath Volume		1.484	21:23
Air Blank		0.000	21:24
Air Blank		0.000	21:26
Subject Sample	# 2	0.073	21:26
Breath Volume		1.207	21:26
Air Blank		0.000	21:27
Control Test		0.083	21:27
Air Blank		0.000	21:28
Diagnostics Check		OK	21:28

Data Download Date/Time: 02/07/2020 08:36

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEARSE,ANGIE,D
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MANGEL,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: MASLIN

DATE/TIME OF TEST: 02/08/2020 00:53:15

Diagnostics Check		OK	00:50	
Air Blank		0.000	00:50	
Control Test		0.083	00:51	
Air Blank		0.000	00:51	
Subject Sample	# 1		00:51	Subject Test Refused
Breath Volume		0.000	00:51	
Air Blank		0.000	00:52	
Control Test		0.083	00:52	
Air Blank		0.000	00:53	
Diagnostics Check		OK	00:53	

Data Download Date/Time: 03/06/2020 04:03

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KING,SHAKAYLA,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,BRANDON
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: DIMARCO

DATE/TIME OF TEST: 02/08/2020 02:09:59

Diagnostics Check		OK	02:07	
Air Blank		0.000	02:07	
Control Test		0.083	02:07	
Air Blank		0.000	02:08	
Subject Sample	# 1		02:08	Subject Test Refused
Breath Volume		0.000	02:08	
Air Blank		0.000	02:09	
Control Test		0.083	02:09	
Air Blank		0.000	02:09	
Diagnostics Check		OK	02:09	

Data Download Date/Time: 03/06/2020 04:03

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIGNORELLI,A,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,B
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: GATES

DATE/TIME OF TEST: 02/08/2020 23:40:41

Diagnostics Check		OK	23:37	
Air Blank		0.000	23:38	
Control Test		0.083	23:38	
Air Blank		0.000	23:39	
Subject Sample	# 1		23:39	Subject Test Refused
Breath Volume		0.000	23:39	
Air Blank		0.000	23:39	
Control Test		0.082	23:40	
Air Blank		0.000	23:40	
Diagnostics Check		OK	23:40	

Data Download Date/Time: 03/06/2020 04:03

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIGNORELLI,A,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: BARCARO,B
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 02/11/2020 00:04:11

Diagnostics Check		OK	23:56
Air Blank		0.000	23:57
Control Test		0.082	23:57
Air Blank		0.000	23:58
Subject Sample	# 2	0.136	23:59
Breath Volume		1.273	23:59
Air Blank		0.000	00:00
Air Blank		0.000	00:01
Subject Sample	# 1	0.134	00:02
Breath Volume		1.628	00:02
Air Blank		0.000	00:03
Control Test		0.082	00:03
Air Blank		0.000	00:04
Diagnostics Check		OK	00:04

Data Download Date/Time: 03/06/2020 04:03

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANDY,KEVIN,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MACIAS,E.H.
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 18:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: DEAN

DATE/TIME OF TEST: 02/12/2020 19:06:38

Diagnostics Check		OK	19:00
Air Blank		0.000	19:00
Control Test		0.083	19:00
Air Blank		0.000	19:01
Subject Sample	# 1	0.229	19:01
Breath Volume		2.238	19:01
Air Blank		0.000	19:02
Air Blank		0.000	19:04
Subject Sample	# 2	0.230	19:05
Breath Volume		2.527	19:05
Air Blank		0.000	19:05
Control Test		0.082	19:06
Air Blank		0.000	19:06
Diagnostics Check		OK	19:06

Data Download Date/Time: 03/06/2020 04:03

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KING,SHAKAYLA,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,BRANDON
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: SWING

DATE/TIME OF TEST: 02/13/2020 04:32:26

Diagnostics Check		OK	04:26
Air Blank		0.000	04:26
Control Test		0.083	04:26
Air Blank		0.000	04:27
Subject Sample	# 1	0.171	04:27
Breath Volume		2.535	04:27
Air Blank		0.000	04:28
Air Blank		0.000	04:30
Subject Sample	# 2	0.165	04:30
Breath Volume		1.964	04:30
Air Blank		0.000	04:31
Control Test		0.081	04:31
Air Blank		0.000	04:32
Diagnostics Check		OK	04:32

Data Download Date/Time: 03/06/2020 04:03

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HISS,J,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: BRASHEARS,E
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 19:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:34

SUBJECT'S LAST NAME: BOSWORTH

DATE/TIME OF TEST: 02/13/2020 20:02:18

Diagnostics Check		OK	19:54
Air Blank		0.000	19:55
Control Test		0.083	19:55
Air Blank		0.000	19:55
Subject Sample	# 1	0.252	19:57
Breath Volume		3.101	19:57
Air Blank		0.000	19:57
Air Blank		0.000	19:59
Subject Sample	# 2	0.250	20:00
Breath Volume		2.023	20:00
Air Blank		0.000	20:01
Control Test		0.082	20:01
Air Blank		0.000	20:02
Diagnostics Check		OK	20:02

Data Download Date/Time: 03/06/2020 04:03

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DEPUTY,T
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,OFFICER
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 19:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: BONECUTTER

DATE/TIME OF TEST: 02/13/2020 20:54:30

Diagnostics Check		OK	20:47
Air Blank		0.000	20:48
Control Test		0.082	20:48
Air Blank		0.000	20:48
Subject Sample	# 1	0.056	20:49
Breath Volume		1.234	20:49
Air Blank		0.000	20:50
Air Blank		0.000	20:51
Subject Sample	# 2	0.052	20:52
Breath Volume		1.246	20:52
Air Blank		0.000	20:53
Control Test		0.083	20:53
Air Blank		0.000	20:54
Diagnostics Check		OK	20:54

Data Download Date/Time: 03/06/2020 04:03

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KING,SHAKAYLA,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKEON,A
ARRESTING OFFICER AGENCY: Fellsmere PD
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: GALLEGOS

DATE/TIME OF TEST: 02/14/2020 03:05:11

Diagnostics Check		OK	02:52
Air Blank		0.000	02:52
Control Test		0.083	02:53
Air Blank		0.000	02:53
Subject Sample	# 1	0.263	02:54
Breath Volume		2.714	02:54
Air Blank		0.000	02:55
Air Blank		0.000	02:57
Subject Sample	# 2	0.221	02:58
Breath Volume		1.242	02:58
Air Blank		0.000	02:59
Air Blank		0.000	03:01
Subject Sample	# 3	0.227	03:03
Breath Volume		1.664	03:03
Air Blank		0.000	03:04
Control Test		0.081	03:04
Air Blank		0.000	03:05
Diagnostics Check		OK	03:05

Data Download Date/Time: 03/06/2020 04:03

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIGNORELLI,A,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: YORK,J
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 20:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:19

SUBJECT'S LAST NAME: CHRISTMAN

DATE/TIME OF TEST: 02/15/2020 21:47:43

Diagnostics Check		OK	21:40
Air Blank		0.000	21:40
Control Test		0.084	21:40
Air Blank		0.000	21:41
Subject Sample	# 1	0.157	21:42
Breath Volume		1.761	21:42
Air Blank		0.000	21:42
Air Blank		0.000	21:44
Subject Sample	# 2	0.176	21:45
Breath Volume		3.996	21:45
Air Blank		0.000	21:46
Control Test		0.083	21:47
Air Blank		0.000	21:47
Diagnostics Check		OK	21:47

Data Download Date/Time: 03/06/2020 04:03

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHILE,C,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: BRASHEARS,E
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 02/15/2020 22:30:08

Diagnostics Check		OK	22:21
Air Blank		0.000	22:22
Control Test		0.083	22:22
Air Blank		0.000	22:22
Subject Sample	# 1	0.190	22:23
Breath Volume		1.269	22:23
Air Blank		0.000	22:24
Air Blank		0.000	22:25
Subject Sample	# 2	0.205	22:28
Breath Volume		2.226	22:28
Air Blank		0.000	22:29
Control Test		0.082	22:29
Air Blank		0.000	22:30
Diagnostics Check		OK	22:30

Data Download Date/Time: 03/06/2020 04:03

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIGNORELLI,A,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: JAWORSKI,K
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 15:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:05

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 02/17/2020 16:38:08

Diagnostics Check		OK	16:29
Air Blank		0.000	16:29
Control Test		0.084	16:30
Air Blank		0.000	16:30
Subject Sample	# 1	0.000	16:33
Breath Volume		1.761	16:33
Air Blank		0.000	16:33
Air Blank		0.000	16:35
Subject Sample	# 2	0.000	16:36
Breath Volume		1.105	16:36
Air Blank		0.000	16:37
Control Test		0.083	16:37
Air Blank		0.000	16:38
Diagnostics Check		OK	16:38

Data Download Date/Time: 03/06/2020 04:03

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIGNORELLI,A,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: SWEENEY,C
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 19:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: LOCKETT

DATE/TIME OF TEST: 02/18/2020 21:02:54

Diagnostics Check		OK	20:55
Air Blank		0.000	20:56
Control Test		0.083	20:56
Air Blank		0.000	20:56
Subject Sample	# 1	0.318	20:57
Breath Volume		1.480	20:57
Air Blank		0.000	20:58
Air Blank		0.000	21:00
Subject Sample	# 2	0.309	21:01
Breath Volume		1.812	21:01
Air Blank		0.000	21:01
Control Test		0.082	21:02
Air Blank		0.000	21:02
Diagnostics Check		OK	21:02

Data Download Date/Time: 03/06/2020 04:03

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KANE,JAMES,G
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: BUFFINGTON,BEN
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 14:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:30

SUBJECT'S LAST NAME: SANTANIELLO

DATE/TIME OF TEST: 02/19/2020 14:54:22

Diagnostics Check		OK	14:50	
Air Blank		0.000	14:51	
Control Test		0.082	14:51	
Air Blank		0.000	14:52	
Subject Sample	# 1		14:52	Interferent Detect
Breath Volume		2.960	14:52	
Air Blank		0.000	14:53	
Control Test		0.082	14:53	
Air Blank		0.000	14:54	
Diagnostics Check		OK	14:54	

Data Download Date/Time: 03/06/2020 04:03

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KANE,JAMES,G
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: BUFFINGTON,BEN
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 14:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:30

SUBJECT'S LAST NAME: SANTANIELLO

DATE/TIME OF TEST: 02/19/2020 14:59:23

Diagnostics Check		OK	14:56	
Air Blank		0.000	14:56	
Control Test		0.082	14:56	
Air Blank		0.000	14:57	
Subject Sample	# 1		14:57	Interferent Detect
Breath Volume		2.261	14:57	
Air Blank		0.000	14:58	
Control Test		0.083	14:58	
Air Blank		0.000	14:59	
Diagnostics Check		OK	14:59	

Data Download Date/Time: 03/06/2020 04:03

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DEPUTY,T
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: WARD,DEPUTY
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 16:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: RESNICK

DATE/TIME OF TEST: 02/19/2020 17:20:45

Diagnostics Check		OK	17:14
Air Blank		0.000	17:14
Control Test		0.083	17:14
Air Blank		0.000	17:15
Subject Sample	# 1	0.000	17:16
Breath Volume		1.917	17:16
Air Blank		0.000	17:16
Air Blank		0.000	17:18
Subject Sample	# 2	0.000	17:19
Breath Volume		1.640	17:19
Air Blank		0.000	17:19
Control Test		0.082	17:20
Air Blank		0.000	17:20
Diagnostics Check		OK	17:20

Data Download Date/Time: 03/06/2020 04:03

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHILE,C,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: SANDERS,B
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: GASPARRO

DATE/TIME OF TEST: 02/19/2020 23:00:20

Diagnostics Check		OK	22:53
Air Blank		0.000	22:53
Control Test		0.083	22:53
Air Blank		0.000	22:54
Subject Sample	# 1	0.200	22:55
Breath Volume		1.324	22:55
Air Blank		0.000	22:55
Air Blank		0.000	22:57
Subject Sample	# 2	0.216	22:58
Breath Volume		2.035	22:58
Air Blank		0.000	22:59
Control Test		0.082	22:59
Air Blank		0.000	23:00
Diagnostics Check		OK	23:00

Data Download Date/Time: 03/06/2020 04:03

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANDY,KEVIN,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: RUIZ,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: PETERS

DATE/TIME OF TEST: 02/20/2020 02:59:31

Diagnostics Check		OK	02:53
Air Blank		0.000	02:53
Control Test		0.082	02:53
Air Blank		0.000	02:54
Subject Sample	# 1	0.150	02:54
Breath Volume		2.710	02:54
Air Blank		0.000	02:55
Air Blank		0.000	02:57
Subject Sample	# 2	0.152	02:57
Breath Volume		1.710	02:57
Air Blank		0.000	02:58
Control Test		0.083	02:58
Air Blank		0.000	02:59
Diagnostics Check		OK	02:59

Data Download Date/Time: 03/06/2020 04:03

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLASSE,RAYMOND,J
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MANGEL,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: MARTOS

DATE/TIME OF TEST: 02/20/2020 03:39:19

Diagnostics Check		OK	03:36	
Air Blank		0.000	03:36	
Control Test		0.082	03:37	
Air Blank		0.000	03:37	
Subject Sample	# 1		03:37	Subject Test Refused
Breath Volume		0.000	03:37	
Air Blank		0.000	03:38	
Control Test		0.082	03:38	
Air Blank		0.000	03:39	
Diagnostics Check		OK	03:39	

Data Download Date/Time: 03/06/2020 04:03

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUDOLPH,B,D
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: SANDERS,B
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: WEISS

DATE/TIME OF TEST: 02/20/2020 04:36:14

Diagnostics Check		OK	04:29
Air Blank		0.000	04:30
Control Test		0.083	04:30
Air Blank		0.000	04:30
Subject Sample	# 1	0.177	04:31
Breath Volume		1.554	04:31
Air Blank		0.000	04:31
Air Blank		0.000	04:33
Subject Sample	# 2	0.172	04:34
Breath Volume		1.929	04:34
Air Blank		0.000	04:35
Control Test		0.082	04:35
Air Blank		0.000	04:36
Diagnostics Check		OK	04:36

Data Download Date/Time: 03/06/2020 04:03

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KANE,JAMES,G
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: HOYT,SHAWN
ARRESTING OFFICER AGENCY: Indian River Shores
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 09:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:03

SUBJECT'S LAST NAME: DUTEAU

DATE/TIME OF TEST: 02/20/2020 10:30:31

Diagnostics Check		OK	10:23
Air Blank		0.000	10:23
Control Test		0.082	10:24
Air Blank		0.000	10:24
Subject Sample	# 1	0.236	10:25
Breath Volume		1.765	10:25
Air Blank		0.000	10:26
Air Blank		0.000	10:27
Subject Sample	# 2	0.237	10:28
Breath Volume		2.562	10:28
Air Blank		0.000	10:29
Control Test		0.083	10:29
Air Blank		0.000	10:30
Diagnostics Check		OK	10:30

Data Download Date/Time: 03/06/2020 04:03

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSETT,DONNELLA,L
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: DELEON,TEA
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: VELENTZAS

DATE/TIME OF TEST: 02/20/2020 22:05:04

Diagnostics Check		OK	21:53	
Air Blank		0.000	21:54	
Control Test		0.083	21:54	
Air Blank		0.000	21:54	
Subject Sample	# 1		21:58	No Sample Provided
Breath Volume		0.000	21:58	
Air Blank		0.000	21:58	
Air Blank		0.000	22:00	
Subject Sample	# 2		22:03	No Sample Provided
Breath Volume		0.000	22:03	
Air Blank		0.000	22:04	
Control Test		0.082	22:04	
Air Blank		0.000	22:04	
Diagnostics Check		OK	22:05	

Data Download Date/Time: 03/06/2020 04:03

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSETT,DONNELLA,L
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: DELON,TEA
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: VELENTZAS

DATE/TIME OF TEST: 02/20/2020 22:18:36

Diagnostics Check		OK	22:11
Air Blank		0.000	22:12
Control Test		0.083	22:12
Air Blank		0.000	22:13
Subject Sample	# 1	0.000	22:14
Breath Volume		2.492	22:14
Air Blank		0.000	22:14
Air Blank		0.000	22:16
Subject Sample	# 2	0.000	22:17
Breath Volume		1.414	22:17
Air Blank		0.000	22:17
Control Test		0.081	22:17
Air Blank		0.000	22:18
Diagnostics Check		OK	22:18

Data Download Date/Time: 03/06/2020 04:03

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,DEVIN,C
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: CASTANO,C
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: CAMPOS SOTELO

DATE/TIME OF TEST: 02/21/2020 23:47:21

Diagnostics Check		OK	23:39
Air Blank		0.000	23:39
Control Test		0.084	23:39
Air Blank		0.000	23:40
Subject Sample	# 1	0.197	23:42
Breath Volume		2.632	23:42
Air Blank		0.000	23:43
Air Blank		0.000	23:44
Subject Sample	# 2	0.204	23:45
Breath Volume		2.312	23:45
Air Blank		0.000	23:46
Control Test		0.082	23:46
Air Blank		0.000	23:47
Diagnostics Check		OK	23:47

Data Download Date/Time: 03/06/2020 04:03

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KING,SHAKAYLA,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MANGEL,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: MELTON

DATE/TIME OF TEST: 02/22/2020 00:48:13

Diagnostics Check		OK	00:41
Air Blank		0.000	00:42
Control Test		0.083	00:42
Air Blank		0.000	00:42
Subject Sample	# 1	0.245	00:43
Breath Volume		3.656	00:43
Air Blank		0.000	00:44
Air Blank		0.000	00:46
Subject Sample	# 2	0.229	00:46
Breath Volume		1.316	00:46
Air Blank		0.000	00:47
Control Test		0.082	00:47
Air Blank		0.000	00:48
Diagnostics Check		OK	00:48

Data Download Date/Time: 03/06/2020 04:03

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,DEVIN,C
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: BRASHEARS,E
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 02/22/2020 02:15:41

Diagnostics Check		OK	02:08
Air Blank		0.000	02:09
Control Test		0.083	02:09
Air Blank		0.000	02:10
Subject Sample	# 1	0.122	02:11
Breath Volume		4.316	02:11
Air Blank		0.000	02:11
Air Blank		0.000	02:13
Subject Sample	# 2	0.118	02:14
Breath Volume		3.390	02:14
Air Blank		0.000	02:14
Control Test		0.081	02:15
Air Blank		0.000	02:15
Diagnostics Check		OK	02:15

Data Download Date/Time: 03/06/2020 04:03

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MAEGAN,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: FARINACCI,DYLAN
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 17:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:25

SUBJECT'S LAST NAME: SCAVONE

DATE/TIME OF TEST: 02/22/2020 19:09:49

Diagnostics Check		OK	19:01
Air Blank		0.000	19:01
Control Test		0.083	19:02
Air Blank		0.000	19:02
Subject Sample	# 1	0.146	19:04
Breath Volume		1.445	19:04
Air Blank		0.000	19:05
Air Blank		0.000	19:06
Subject Sample	# 2	0.144	19:08
Breath Volume		1.437	19:08
Air Blank		0.000	19:08
Control Test		0.082	19:09
Air Blank		0.000	19:09
Diagnostics Check		OK	19:09

Data Download Date/Time: 03/06/2020 04:03

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,DEVIN,C
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: FARINACCI,DYLAN
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: WILDER

DATE/TIME OF TEST: 02/23/2020 00:24:37

Diagnostics Check		OK	00:18
Air Blank		0.000	00:18
Control Test		0.083	00:18
Air Blank		0.000	00:19
Subject Sample	# 1	0.000	00:20
Breath Volume		3.289	00:20
Air Blank		0.000	00:20
Air Blank		0.000	00:22
Subject Sample	# 2	0.000	00:23
Breath Volume		2.945	00:23
Air Blank		0.000	00:23
Control Test		0.083	00:24
Air Blank		0.000	00:24
Diagnostics Check		OK	00:24

Data Download Date/Time: 03/06/2020 04:03

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KING,SHAKAYLA,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MANGEL,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: MOHR

DATE/TIME OF TEST: 02/23/2020 03:16:11

Diagnostics Check		OK	03:09
Air Blank		0.000	03:10
Control Test		0.082	03:10
Air Blank		0.000	03:11
Subject Sample	# 1	0.201	03:11
Breath Volume		1.347	03:11
Air Blank		0.000	03:12
Air Blank		0.000	03:14
Subject Sample	# 2	0.214	03:14
Breath Volume		2.371	03:14
Air Blank		0.000	03:15
Control Test		0.082	03:15
Air Blank		0.000	03:16
Diagnostics Check		OK	03:16

Data Download Date/Time: 03/06/2020 04:03

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,DEVIN,C
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,BRANDON
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: LEMAY

DATE/TIME OF TEST: 02/23/2020 03:28:46

Diagnostics Check		OK	03:25	
Air Blank		0.000	03:26	
Control Test		0.082	03:26	
Air Blank		0.000	03:27	
Subject Sample	# 1		03:27	Subject Test Refused
Breath Volume		0.000	03:27	
Air Blank		0.000	03:27	
Control Test		0.084	03:28	
Air Blank		0.000	03:28	
Diagnostics Check		OK	03:28	

Data Download Date/Time: 03/06/2020 04:04

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIGNORELLI,A,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,B
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 21:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: MORE

DATE/TIME OF TEST: 02/23/2020 22:01:08

Diagnostics Check		OK	21:52
Air Blank		0.000	21:52
Control Test		0.083	21:53
Air Blank		0.000	21:53
Subject Sample	# 1	0.147	21:56
Breath Volume		1.398	21:56
Air Blank		0.000	21:57
Air Blank		0.000	21:58
Subject Sample	# 2	0.147	21:59
Breath Volume		1.363	21:59
Air Blank		0.000	22:00
Control Test		0.083	22:00
Air Blank		0.000	22:01
Diagnostics Check		OK	22:01

Data Download Date/Time: 03/06/2020 04:04

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,BRIAN,P
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: GREEN,BRIAN
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 21:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: EMLET

DATE/TIME OF TEST: 02/23/2020 23:57:31

Diagnostics Check		OK	23:51
Air Blank		0.000	23:51
Control Test		0.083	23:52
Air Blank		0.000	23:52
Subject Sample	# 1	0.000	23:53
Breath Volume		2.910	23:53
Air Blank		0.000	23:53
Air Blank		0.000	23:55
Subject Sample	# 2	0.000	23:56
Breath Volume		2.175	23:56
Air Blank		0.000	23:56
Control Test		0.083	23:56
Air Blank		0.000	23:57
Diagnostics Check		OK	23:57

Data Download Date/Time: 03/06/2020 04:04

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIGNORELLI,A,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MARTY,K
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 16:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: KREISEL

DATE/TIME OF TEST: 02/25/2020 18:57:29

Diagnostics Check		OK	18:49
Air Blank		0.000	18:50
Control Test		0.083	18:50
Air Blank		0.000	18:50
Subject Sample	# 1	0.285	18:52
Breath Volume		1.335	18:52
Air Blank		0.000	18:52
Air Blank		0.000	18:54
Subject Sample	# 2	0.273	18:55
Breath Volume		1.800	18:55
Air Blank		0.000	18:56
Control Test		0.082	18:56
Air Blank		0.000	18:57
Diagnostics Check		OK	18:57

Data Download Date/Time: 03/06/2020 04:04

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOOD,T,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: DAVIS,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 08:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:35

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 02/26/2020 09:04:11

Diagnostics Check		OK	08:57
Air Blank		0.000	08:57
Control Test		0.083	08:58
Air Blank		0.000	08:58
Subject Sample	# 1	0.264	08:59
Breath Volume		2.460	08:59
Air Blank		0.000	08:59
Air Blank		0.000	09:01
Subject Sample	# 2	0.263	09:02
Breath Volume		2.605	09:02
Air Blank		0.000	09:03
Control Test		0.083	09:03
Air Blank		0.000	09:04
Diagnostics Check		OK	09:04

Data Download Date/Time: 03/06/2020 04:04

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHILE,C,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MANGEL,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 19:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: CIVILLE

DATE/TIME OF TEST: 02/26/2020 21:11:43

Diagnostics Check		OK	21:05
Air Blank		0.000	21:06
Control Test		0.083	21:06
Air Blank		0.000	21:07
Subject Sample	# 1	0.000	21:07
Breath Volume		1.457	21:07
Air Blank		0.000	21:08
Air Blank		0.000	21:10
Subject Sample	# 2	0.000	21:10
Breath Volume		1.316	21:10
Air Blank		0.000	21:10
Control Test		0.083	21:11
Air Blank		0.000	21:11
Diagnostics Check		OK	21:11

Data Download Date/Time: 03/06/2020 04:04

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHILE,C,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: FERGUSON,T
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 20:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: LEVITT

DATE/TIME OF TEST: 02/26/2020 22:47:16

Diagnostics Check		OK	22:34	
Air Blank		0.000	22:34	
Control Test		0.082	22:35	
Air Blank		0.000	22:35	
Subject Sample	# 1		22:38	No Sample Provided
Breath Volume		0.000	22:38	
Air Blank		0.000	22:39	
Air Blank		0.000	22:41	
Subject Sample	# 2	0.056	22:42	
Breath Volume		1.484	22:42	
Air Blank		0.000	22:43	
Air Blank		0.000	22:45	
Subject Sample	# 3	0.057	22:45	
Breath Volume		1.199	22:45	
Air Blank		0.000	22:46	
Control Test		0.083	22:46	
Air Blank		0.000	22:47	
Diagnostics Check		OK	22:47	

Data Download Date/Time: 03/06/2020 04:04

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANDY,KEVIN,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: SANDERS,B
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: MOYLAN

DATE/TIME OF TEST: 02/27/2020 02:12:53

Diagnostics Check		OK	02:09	
Air Blank		0.000	02:10	
Control Test		0.084	02:10	
Air Blank		0.000	02:10	
Subject Sample	# 1		02:11	Subject Test Refused
Breath Volume		0.000	02:11	
Air Blank		0.000	02:11	
Control Test		0.083	02:12	
Air Blank		0.000	02:12	
Diagnostics Check		OK	02:12	

Data Download Date/Time: 03/06/2020 04:04

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHILE,C,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,B
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: ROSEWARNE

DATE/TIME OF TEST: 02/27/2020 22:58:15

Diagnostics Check		OK	22:51
Air Blank		0.000	22:51
Control Test		0.083	22:52
Air Blank		0.000	22:52
Subject Sample	# 1	0.161	22:53
Breath Volume		1.468	22:53
Air Blank		0.000	22:54
Air Blank		0.000	22:56
Subject Sample	# 2	0.156	22:56
Breath Volume		1.769	22:56
Air Blank		0.000	22:57
Control Test		0.083	22:57
Air Blank		0.000	22:58
Diagnostics Check		OK	22:58

Data Download Date/Time: 03/06/2020 04:04

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLASSE,RAYMOND,J
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: SZOTT,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: VALENTINO

DATE/TIME OF TEST: 02/28/2020 00:37:29

Diagnostics Check		OK	00:34	
Air Blank		0.000	00:35	
Control Test		0.083	00:35	
Air Blank		0.000	00:35	
Subject Sample	# 1		00:36	Subject Test Refused
Breath Volume		0.000	00:36	
Air Blank		0.000	00:36	
Control Test		0.083	00:36	
Air Blank		0.000	00:37	
Diagnostics Check		OK	00:37	

Data Download Date/Time: 03/06/2020 04:04

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUDOLPH,B,D
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,B
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: WILKIE

DATE/TIME OF TEST: 02/28/2020 03:28:52

Diagnostics Check		OK	03:25	
Air Blank		0.000	03:26	
Control Test		0.083	03:26	
Air Blank		0.000	03:27	
Subject Sample	# 1		03:27	Subject Test Refused
Breath Volume		0.000	03:27	
Air Blank		0.000	03:27	
Control Test		0.083	03:28	
Air Blank		0.000	03:28	
Diagnostics Check		OK	03:28	

Data Download Date/Time: 03/06/2020 04:04

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARROQUIN,RICHARD,J
BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)
ARRESTING OFFICER: MARROQUIN,RICHARD
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 19:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:37

SUBJECT'S LAST NAME: ZIMMER

DATE/TIME OF TEST: 02/28/2020 20:37:37

Diagnostics Check		OK	20:34	
Air Blank		0.000	20:35	
Control Test		0.084	20:35	
Air Blank		0.000	20:36	
Subject Sample	# 1		20:36	Subject Test Refused
Breath Volume		0.000	20:36	
Air Blank		0.000	20:36	
Control Test		0.084	20:37	
Air Blank		0.000	20:37	
Diagnostics Check		OK	20:37	

Data Download Date/Time: 03/06/2020 04:04

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIGNORELLI,A,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: BRANDS,M
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:43

SUBJECT'S LAST NAME: TYRE

DATE/TIME OF TEST: 02/28/2020 23:10:51

Diagnostics Check		OK	23:03
Air Blank		0.000	23:04
Control Test		0.084	23:04
Air Blank		0.000	23:05
Subject Sample	# 1	0.041	23:06
Breath Volume		1.148	23:06
Air Blank		0.000	23:06
Air Blank		0.000	23:08
Subject Sample	# 2	0.039	23:09
Breath Volume		1.277	23:09
Air Blank		0.000	23:09
Control Test		0.084	23:10
Air Blank		0.000	23:10
Diagnostics Check		OK	23:10

Data Download Date/Time: 03/06/2020 04:04

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,DEVIN,C
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MANGEL,MELINDA
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 02/29/2020 02:20:52

Diagnostics Check		OK	02:18	
Air Blank		0.000	02:18	
Control Test		0.084	02:18	
Air Blank		0.000	02:19	
Subject Sample	# 1		02:19	Subject Test Refused
Breath Volume		0.000	02:19	
Air Blank		0.000	02:19	
Control Test		0.084	02:20	
Air Blank		0.000	02:20	
Diagnostics Check		OK	02:20	

Data Download Date/Time: 03/06/2020 04:04

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHILE,C,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: BRANDES,B
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 20:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: HUFF

DATE/TIME OF TEST: 02/29/2020 21:46:18

Diagnostics Check		OK	21:43	
Air Blank		0.000	21:43	
Control Test		0.084	21:44	
Air Blank		0.000	21:44	
Subject Sample	# 1		21:44	Subject Test Refused
Breath Volume		0.000	21:44	
Air Blank		0.000	21:45	
Control Test		0.084	21:45	
Air Blank		0.000	21:46	
Diagnostics Check		OK	21:46	

Data Download Date/Time: 03/06/2020 04:04

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,JUAN,
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: VASQUEZ,BRYANT
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 02/02/2020 00:38:26

Diagnostics Check		OK	00:35	
Air Blank		0.000	00:35	
Control Test		0.081	00:36	
Air Blank		0.000	00:36	
Subject Sample	# 1	0.203	00:36	
Breath Volume		1.191	00:36	
Air Blank			00:37	Improper Sample
Air Blank		0.000	00:37	
Control Test		0.081	00:37	
Air Blank		0.000	00:38	
Diagnostics Check		OK	00:38	

Data Download Date/Time: 02/03/2020 09:54

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,JUAN,
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: VASQUEZ,JUAN
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 02/02/2020 00:48:42

Diagnostics Check		OK	00:42
Air Blank		0.000	00:42
Control Test		0.080	00:43
Air Blank		0.000	00:43
Subject Sample	# 1	0.199	00:44
Breath Volume		1.460	00:44
Air Blank		0.000	00:44
Air Blank		0.000	00:46
Subject Sample	# 2	0.200	00:47
Breath Volume		1.363	00:47
Air Blank		0.000	00:47
Control Test		0.080	00:48
Air Blank		0.000	00:48
Diagnostics Check		OK	00:48

Data Download Date/Time: 02/03/2020 09:54

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BALDERAS,GABRIEL,
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: VASQUEZ,BRYANT
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 02/02/2020 01:01:51

Diagnostics Check		OK	00:55
Air Blank		0.000	00:55
Control Test		0.080	00:56
Air Blank		0.000	00:56
Subject Sample	# 1	0.190	00:57
Breath Volume		1.390	00:57
Air Blank		0.000	00:58
Air Blank		0.000	00:59
Subject Sample	# 2	0.191	01:00
Breath Volume		1.511	01:00
Air Blank		0.000	01:00
Control Test		0.080	01:01
Air Blank		0.000	01:01
Diagnostics Check		OK	01:01

Data Download Date/Time: 02/03/2020 09:54

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HISCOCK,MANUELA,E
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: STRUBE,SAMUEL
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:16

SUBJECT'S LAST NAME: YATES

DATE/TIME OF TEST: 02/04/2020 22:42:20

Diagnostics Check		OK	22:36
Air Blank		0.000	22:36
Control Test		0.081	22:37
Air Blank		0.000	22:37
Subject Sample	# 1	0.000	22:38
Breath Volume		2.085	22:38
Air Blank		0.000	22:38
Air Blank		0.000	22:40
Subject Sample	# 2	0.000	22:40
Breath Volume		2.039	22:40
Air Blank		0.000	22:41
Control Test		0.081	22:41
Air Blank		0.000	22:42
Diagnostics Check		OK	22:42

Data Download Date/Time: 02/26/2020 00:40

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HISCOCK,MANUELA,E
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: STRUBE,SAMUEL
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:16

SUBJECT'S LAST NAME: YATES

DATE/TIME OF TEST: 02/04/2020 22:53:08

Diagnostics Check		OK	22:47
Air Blank		0.000	22:47
Control Test		0.081	22:47
Air Blank		0.000	22:48
Subject Sample	# 1	0.000	22:48
Breath Volume		2.234	22:48
Air Blank		0.000	22:49
Air Blank		0.000	22:51
Subject Sample	# 2	0.000	22:51
Breath Volume		1.894	22:51
Air Blank		0.000	22:52
Control Test		0.080	22:52
Air Blank		0.000	22:53
Diagnostics Check		OK	22:53

Data Download Date/Time: 02/26/2020 00:40

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VEGA,JASON,
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: RIOS,DANIELLE
ARRESTING OFFICER AGENCY: Arcadia PD
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 06:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:42

SUBJECT'S LAST NAME: PIERCE

DATE/TIME OF TEST: 02/08/2020 07:11:00

Diagnostics Check		OK	07:04
Air Blank		0.000	07:04
Control Test		0.081	07:05
Air Blank		0.000	07:05
Subject Sample	# 1	0.239	07:06
Breath Volume		2.640	07:06
Air Blank		0.000	07:06
Air Blank		0.000	07:08
Subject Sample	# 2	0.240	07:09
Breath Volume		1.371	07:09
Air Blank		0.000	07:10
Control Test		0.080	07:10
Air Blank		0.000	07:10
Diagnostics Check		OK	07:11

Data Download Date/Time: 02/26/2020 00:40

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VEGA,JASON,
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: RIOS,DANIELLE
ARRESTING OFFICER AGENCY: Arcadia PD
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 06:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:42

SUBJECT'S LAST NAME: PIERCE

DATE/TIME OF TEST: 02/08/2020 07:20:15

Diagnostics Check		OK	07:14
Air Blank		0.000	07:14
Control Test		0.080	07:14
Air Blank		0.000	07:15
Subject Sample	# 1	0.243	07:15
Breath Volume		2.527	07:15
Air Blank		0.000	07:16
Air Blank		0.000	07:18
Subject Sample	# 2	0.238	07:18
Breath Volume		1.968	07:18
Air Blank		0.000	07:19
Control Test		0.080	07:19
Air Blank		0.000	07:20
Diagnostics Check		OK	07:20

Data Download Date/Time: 02/26/2020 00:40

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DITTMAN,LAVERNE,T
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: CARROLL,STEVEN
ARRESTING OFFICER AGENCY: Arcadia PD
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: BARAJAS

DATE/TIME OF TEST: 02/09/2020 00:52:02

Diagnostics Check		OK	00:45
Air Blank		0.000	00:46
Control Test		0.081	00:46
Air Blank		0.000	00:46
Subject Sample	# 1	0.274	00:47
Breath Volume		1.593	00:47
Air Blank		0.000	00:48
Air Blank		0.000	00:49
Subject Sample	# 2	0.274	00:50
Breath Volume		1.562	00:50
Air Blank		0.000	00:51
Control Test		0.080	00:51
Air Blank		0.000	00:51
Diagnostics Check		OK	00:52

Data Download Date/Time: 02/26/2020 00:40

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BALDERAS,GABRIEL,
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: VASQUEZ,BRYANT
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: CLEMENTS

DATE/TIME OF TEST: 02/16/2020 04:59:29

Diagnostics Check		OK	04:52
Air Blank		0.000	04:53
Control Test		0.081	04:53
Air Blank		0.000	04:54
Subject Sample	# 1	0.056	04:54
Breath Volume		1.589	04:54
Air Blank		0.000	04:55
Air Blank		0.000	04:57
Subject Sample	# 2	0.053	04:58
Breath Volume		1.515	04:58
Air Blank		0.000	04:58
Control Test		0.080	04:58
Air Blank		0.000	04:59
Diagnostics Check		OK	04:59

Data Download Date/Time: 02/26/2020 00:40

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DITTMAN,LAVERNE,T
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: STRUBE,SAMUEL
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 02/22/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: BERMUDES DIAZ

DATE/TIME OF TEST: 02/23/2020 01:21:33

Diagnostics Check		OK	01:14
Air Blank		0.000	01:15
Control Test		0.081	01:15
Air Blank		0.000	01:16
Subject Sample	# 1	0.134	01:16
Breath Volume		1.539	01:16
Air Blank		0.000	01:17
Air Blank		0.000	01:19
Subject Sample	# 2	0.142	01:20
Breath Volume		1.425	01:20
Air Blank		0.000	01:20
Control Test		0.081	01:20
Air Blank		0.000	01:21
Diagnostics Check		OK	01:21

Data Download Date/Time: 02/26/2020 00:40

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELLFLOWER,MICHAEL,D
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: VASQUEZ,BRYANT
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 02/22/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: CHO CHUNG HING

DATE/TIME OF TEST: 02/25/2020 23:58:11

Diagnostics Check		OK	23:46	
Air Blank		0.000	23:47	
Control Test		0.081	23:47	
Air Blank		0.000	23:47	
Subject Sample	# 1	0.141	23:51	Volume Not Met
Breath Volume		0.363	23:51	
Air Blank		0.000	23:51	
Air Blank		0.000	23:53	
Subject Sample	# 2	0.146	23:56	Volume Not Met
Breath Volume		0.207	23:56	
Air Blank		0.000	23:57	
Control Test		0.081	23:57	
Air Blank		0.000	23:58	
Diagnostics Check		OK	23:58	

Data Download Date/Time: 02/26/2020 00:40

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,NICKKI,N
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: ROWELL,J
ARRESTING OFFICER AGENCY: Venice PD
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: SHERMAN

DATE/TIME OF TEST: 02/02/2020 04:17:47

Diagnostics Check		OK	04:08
Air Blank		0.000	04:08
Control Test		0.081	04:09
Air Blank		0.000	04:09
Subject Sample	# 1	0.172	04:10
Breath Volume		1.746	04:10
Air Blank		0.000	04:10
Air Blank		0.000	04:12
Subject Sample	# 2	0.193	04:13
Breath Volume		1.382	04:13
Air Blank		0.000	04:14
Air Blank		0.000	04:15
Subject Sample	# 3	0.170	04:16
Breath Volume		1.863	04:16
Air Blank		0.000	04:16
Control Test		0.079	04:17
Air Blank		0.000	04:17
Diagnostics Check		OK	04:17

Data Download Date/Time: 02/26/2020 06:12

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BOWERS,JOSEPH
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: NERI

DATE/TIME OF TEST: 02/02/2020 22:54:56

Diagnostics Check		OK	22:50	
Air Blank		0.000	22:51	
Control Test		0.080	22:51	
Air Blank		0.000	22:51	
Subject Sample	# 1		22:53	Subject Test Refused
Breath Volume		0.000	22:53	
Air Blank		0.000	22:54	
Control Test		0.079	22:54	
Air Blank		0.000	22:54	
Diagnostics Check		OK	22:54	

Data Download Date/Time: 02/26/2020 06:12

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,J,S
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: GRIFFEN,A
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: CHARMLEY

DATE/TIME OF TEST: 02/03/2020 02:02:52

Diagnostics Check		OK	01:59	
Air Blank		0.000	01:59	
Control Test		0.080	02:00	
Air Blank		0.000	02:00	
Subject Sample	# 1	0.285	02:01	
Breath Volume		3.117	02:01	
Air Blank			02:01	RFI Detect
Air Blank		0.000	02:01	
Control Test		0.080	02:02	
Air Blank		0.000	02:02	
Diagnostics Check		OK	02:02	

Data Download Date/Time: 02/26/2020 06:12

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,J,S
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: GRIFFEN,A
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: CHARMLEY

DATE/TIME OF TEST: 02/03/2020 02:10:23

Diagnostics Check		OK	02:04
Air Blank		0.000	02:04
Control Test		0.078	02:04
Air Blank		0.000	02:05
Subject Sample	# 1	0.254	02:05
Breath Volume		2.390	02:05
Air Blank		0.000	02:06
Air Blank		0.000	02:08
Subject Sample	# 2	0.245	02:08
Breath Volume		1.804	02:08
Air Blank		0.000	02:09
Control Test		0.076	02:09
Air Blank		0.000	02:10
Diagnostics Check		OK	02:10

Data Download Date/Time: 02/26/2020 06:12

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,DANIEL,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: YODER,DEP.
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: BJORGE

DATE/TIME OF TEST: 02/03/2020 03:07:22

Diagnostics Check		OK	03:01
Air Blank		0.000	03:01
Control Test		0.080	03:02
Air Blank		0.000	03:02
Subject Sample	# 1	0.000	03:03
Breath Volume		2.445	03:03
Air Blank		0.000	03:03
Air Blank		0.000	03:05
Subject Sample	# 2	0.000	03:05
Breath Volume		2.617	03:05
Air Blank		0.000	03:06
Control Test		0.079	03:06
Air Blank		0.000	03:07
Diagnostics Check		OK	03:07

Data Download Date/Time: 02/26/2020 06:12

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,DANIEL,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: COKER,OFC
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: CHRISTOLIAS

DATE/TIME OF TEST: 02/03/2020 04:08:03

Diagnostics Check		OK	04:02
Air Blank		0.000	04:02
Control Test		0.081	04:02
Air Blank		0.000	04:03
Subject Sample	# 1	0.000	04:03
Breath Volume		2.441	04:03
Air Blank		0.000	04:04
Air Blank		0.000	04:06
Subject Sample	# 2	0.000	04:06
Breath Volume		1.562	04:06
Air Blank		0.000	04:07
Control Test		0.079	04:07
Air Blank		0.000	04:07
Diagnostics Check		OK	04:08

Data Download Date/Time: 02/26/2020 06:12

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BOWERS,JOSEPH
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: SMUCKER

DATE/TIME OF TEST: 02/07/2020 03:41:28

Diagnostics Check		OK	03:35
Air Blank		0.000	03:35
Control Test		0.079	03:36
Air Blank		0.000	03:36
Subject Sample	# 1	0.119	03:37
Breath Volume		2.640	03:37
Air Blank		0.000	03:37
Air Blank		0.000	03:39
Subject Sample	# 2	0.121	03:39
Breath Volume		3.695	03:39
Air Blank		0.000	03:40
Control Test		0.077	03:40
Air Blank		0.000	03:41
Diagnostics Check		OK	03:41

Data Download Date/Time: 02/26/2020 06:12

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEPUTY,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 02/07/2020 23:21:09

Diagnostics Check		OK	23:15
Air Blank		0.000	23:15
Control Test		0.078	23:15
Air Blank		0.000	23:16
Subject Sample	# 1	0.096	23:16
Breath Volume		2.425	23:16
Air Blank		0.000	23:17
Air Blank		0.000	23:19
Subject Sample	# 2	0.094	23:19
Breath Volume		1.800	23:19
Air Blank		0.000	23:20
Control Test		0.079	23:20
Air Blank		0.000	23:21
Diagnostics Check		OK	23:21

Data Download Date/Time: 02/26/2020 06:12

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEPUTY,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SANDERS,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: ESGRO

DATE/TIME OF TEST: 02/08/2020 00:22:13

Diagnostics Check		OK	00:15
Air Blank		0.000	00:16
Control Test		0.080	00:16
Air Blank		0.000	00:16
Subject Sample	# 1	0.016	00:17
Breath Volume		1.746	00:17
Air Blank		0.000	00:18
Air Blank		0.000	00:20
Subject Sample	# 2	0.015	00:20
Breath Volume		1.800	00:20
Air Blank		0.000	00:21
Control Test		0.079	00:21
Air Blank		0.000	00:22
Diagnostics Check		OK	00:22

Data Download Date/Time: 02/26/2020 06:12

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAMORS,JASON,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BELL,D
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: RAUMI

DATE/TIME OF TEST: 02/08/2020 02:58:12

Diagnostics Check		OK	02:53	
Air Blank		0.000	02:54	
Control Test		0.080	02:54	
Air Blank		0.000	02:55	
Subject Sample	# 1		02:56	Subject Test Refused
Breath Volume		0.121	02:56	
Air Blank		0.000	02:57	
Control Test		0.080	02:57	
Air Blank		0.000	02:58	
Diagnostics Check		OK	02:58	

Data Download Date/Time: 02/26/2020 06:12

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEPUTY,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: KOCH

DATE/TIME OF TEST: 02/08/2020 04:26:45

Diagnostics Check		OK	04:20
Air Blank		0.000	04:21
Control Test		0.081	04:21
Air Blank		0.000	04:22
Subject Sample	# 1	0.128	04:22
Breath Volume		2.917	04:22
Air Blank		0.000	04:22
Air Blank		0.000	04:24
Subject Sample	# 2	0.127	04:25
Breath Volume		3.039	04:25
Air Blank		0.000	04:25
Control Test		0.080	04:26
Air Blank		0.000	04:26
Diagnostics Check		OK	04:26

Data Download Date/Time: 02/26/2020 06:12

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DEP,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SANDERS,DEP J
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: BALTIERRA PARRA

DATE/TIME OF TEST: 02/08/2020 23:15:46

Diagnostics Check		OK	23:09
Air Blank		0.000	23:10
Control Test		0.080	23:10
Air Blank		0.000	23:11
Subject Sample	# 1	0.000	23:11
Breath Volume		1.656	23:11
Air Blank		0.000	23:12
Air Blank		0.000	23:14
Subject Sample	# 2	0.000	23:14
Breath Volume		1.113	23:14
Air Blank		0.000	23:14
Control Test		0.080	23:15
Air Blank		0.000	23:15
Diagnostics Check		OK	23:15

Data Download Date/Time: 02/26/2020 06:12

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAMORS,JASON,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BELL,D
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:02

SUBJECT'S LAST NAME: KOLESNIKOV

DATE/TIME OF TEST: 02/09/2020 20:35:52

Diagnostics Check		OK	20:29
Air Blank		0.000	20:30
Control Test		0.080	20:30
Air Blank		0.000	20:30
Subject Sample	# 1	0.274	20:31
Breath Volume		2.007	20:31
Air Blank		0.000	20:31
Air Blank		0.000	20:33
Subject Sample	# 2	0.275	20:34
Breath Volume		2.148	20:34
Air Blank		0.000	20:34
Control Test		0.079	20:35
Air Blank		0.000	20:35
Diagnostics Check		OK	20:35

Data Download Date/Time: 02/26/2020 06:12

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEPUTY,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BELL,DUSTIN
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: SPITALE

DATE/TIME OF TEST: 02/10/2020 00:19:37

Diagnostics Check		OK	00:13
Air Blank		0.000	00:14
Control Test		0.080	00:14
Air Blank		0.000	00:14
Subject Sample	# 1	0.100	00:15
Breath Volume		2.300	00:15
Air Blank		0.000	00:15
Air Blank		0.000	00:17
Subject Sample	# 2	0.098	00:18
Breath Volume		2.117	00:18
Air Blank		0.000	00:18
Control Test		0.079	00:19
Air Blank		0.000	00:19
Diagnostics Check		OK	00:19

Data Download Date/Time: 02/26/2020 06:12

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEPUTY,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: HOWELL

DATE/TIME OF TEST: 02/12/2020 22:21:24

Diagnostics Check	OK	22:19	
Air Blank	0.000	22:20	
Control Test	0.079	22:20	
Air Blank		22:20	RFI Detect
Air Blank	0.000	22:21	

Data Download Date/Time: 02/26/2020 06:12

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEPUTY,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: HOWELL

DATE/TIME OF TEST: 02/12/2020 22:29:05

Diagnostics Check		OK	22:23
Air Blank		0.000	22:23
Control Test		0.079	22:23
Air Blank		0.000	22:24
Subject Sample	# 1	0.149	22:24
Breath Volume		2.714	22:24
Air Blank		0.000	22:25
Air Blank		0.000	22:27
Subject Sample	# 2	0.145	22:27
Breath Volume		2.710	22:27
Air Blank		0.000	22:28
Control Test		0.078	22:28
Air Blank		0.000	22:28
Diagnostics Check		OK	22:29

Data Download Date/Time: 02/26/2020 06:12

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,DANIEL,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FOLADVANDI,OFC
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: CULP

DATE/TIME OF TEST: 02/15/2020 23:31:07

Diagnostics Check		OK	23:25
Air Blank		0.000	23:25
Control Test		0.080	23:26
Air Blank		0.000	23:26
Subject Sample	# 1	0.000	23:26
Breath Volume		3.839	23:26
Air Blank		0.000	23:27
Air Blank		0.000	23:29
Subject Sample	# 2	0.000	23:29
Breath Volume		3.792	23:29
Air Blank		0.000	23:30
Control Test		0.080	23:30
Air Blank		0.000	23:31
Diagnostics Check		OK	23:31

Data Download Date/Time: 02/26/2020 06:12

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLYNN,M,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MARTIN,K
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 02/17/2020 04:13:40

Diagnostics Check		OK	04:05
Air Blank		0.000	04:06
Control Test		0.079	04:06
Air Blank		0.000	04:07
Subject Sample	# 1	0.135	04:09
Breath Volume		1.117	04:09
Air Blank		0.000	04:09
Air Blank		0.000	04:11
Subject Sample	# 2	0.150	04:12
Breath Volume		1.828	04:12
Air Blank		0.000	04:12
Control Test		0.077	04:13
Air Blank		0.000	04:13
Diagnostics Check		OK	04:13

Data Download Date/Time: 02/26/2020 06:12

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DEP,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BELL,DEP D
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: RYAN

DATE/TIME OF TEST: 02/18/2020 00:16:15

Diagnostics Check		OK	00:10
Air Blank		0.000	00:10
Control Test		0.079	00:10
Air Blank		0.000	00:11
Subject Sample	# 1	0.105	00:11
Breath Volume		1.472	00:11
Air Blank		0.000	00:12
Air Blank		0.000	00:14
Subject Sample	# 2	0.101	00:14
Breath Volume		1.128	00:14
Air Blank		0.000	00:15
Control Test		0.079	00:15
Air Blank		0.000	00:16
Diagnostics Check		OK	00:16

Data Download Date/Time: 02/26/2020 06:12

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN,SCOTT,H
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: GUZMAN,SCOTT
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 10:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:31

SUBJECT'S LAST NAME: CSENTERI

DATE/TIME OF TEST: 02/18/2020 11:59:44

Diagnostics Check		OK	11:53
Air Blank		0.000	11:54
Control Test		0.080	11:54
Air Blank		0.000	11:54
Subject Sample	# 1	0.211	11:55
Breath Volume		2.734	11:55
Air Blank		0.000	11:55
Air Blank		0.000	11:57
Subject Sample	# 2	0.209	11:58
Breath Volume		3.457	11:58
Air Blank		0.000	11:58
Control Test		0.079	11:59
Air Blank		0.000	11:59
Diagnostics Check		OK	11:59

Data Download Date/Time: 02/26/2020 06:12

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAY,JOHN,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SALSMAN,OFFICER
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: DANSEREAU

DATE/TIME OF TEST: 02/18/2020 21:37:10

Diagnostics Check		OK	21:30
Air Blank		0.000	21:30
Control Test		0.079	21:31
Air Blank		0.000	21:31
Subject Sample	# 1	0.208	21:32
Breath Volume		1.859	21:32
Air Blank		0.000	21:33
Air Blank		0.000	21:35
Subject Sample	# 2	0.214	21:35
Breath Volume		1.972	21:35
Air Blank		0.000	21:36
Control Test		0.078	21:36
Air Blank		0.000	21:37
Diagnostics Check		OK	21:37

Data Download Date/Time: 02/26/2020 06:12

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANTWELL,J,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: CANTWELL,J
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: MONTEFORTE

DATE/TIME OF TEST: 02/19/2020 03:05:16

Diagnostics Check		OK	02:58
Air Blank		0.000	02:59
Control Test		0.080	02:59
Air Blank		0.000	03:00
Subject Sample	# 1	0.149	03:00
Breath Volume		3.378	03:00
Air Blank		0.000	03:01
Air Blank		0.000	03:03
Subject Sample	# 2	0.148	03:03
Breath Volume		3.562	03:03
Air Blank		0.000	03:04
Control Test		0.080	03:04
Air Blank		0.000	03:05
Diagnostics Check		OK	03:05

Data Download Date/Time: 02/26/2020 06:12

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ECKERT,GABRIEL,B
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: ECKERT,GABRIEL
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: HASCAKOVA

DATE/TIME OF TEST: 02/20/2020 23:58:39

Diagnostics Check		OK	23:44	
Air Blank		0.000	23:44	
Control Test		0.079	23:45	
Air Blank		0.000	23:45	
Subject Sample	# 1		23:48	No Sample Provided
Breath Volume		0.000	23:48	
Air Blank		0.000	23:49	
Air Blank		0.000	23:51	
Subject Sample	# 2	0.217	23:51	
Breath Volume		2.363	23:51	
Air Blank		0.000	23:52	
Air Blank		0.000	23:54	
Subject Sample	# 3	0.200	23:56	
Breath Volume		1.425	23:56	
Air Blank		0.000	23:57	
Control Test		0.079	23:58	
Air Blank		0.000	23:58	
Diagnostics Check		OK	23:58	

Data Download Date/Time: 02/26/2020 06:12

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,FRANK EDWARD,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: PIERRE LOUIS,KENOL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 02/21/2020 06:12:38

Diagnostics Check		OK	06:05
Air Blank		0.000	06:06
Control Test		0.079	06:06
Air Blank		0.000	06:07
Subject Sample	# 1	0.158	06:07
Breath Volume		1.265	06:07
Air Blank		0.000	06:08
Air Blank		0.000	06:10
Subject Sample	# 2	0.157	06:11
Breath Volume		2.160	06:11
Air Blank		0.000	06:11
Control Test		0.077	06:12
Air Blank		0.000	06:12
Diagnostics Check		OK	06:12

Data Download Date/Time: 02/26/2020 06:12

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEPUTY,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEPUTY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: PETROSKY

DATE/TIME OF TEST: 02/22/2020 04:22:35

Diagnostics Check		OK	04:15
Air Blank		0.000	04:16
Control Test		0.081	04:16
Air Blank		0.000	04:17
Subject Sample	# 1	0.220	04:17
Breath Volume		3.179	04:17
Air Blank		0.000	04:18
Air Blank		0.000	04:20
Subject Sample	# 2	0.221	04:21
Breath Volume		2.625	04:21
Air Blank		0.000	04:21
Control Test		0.080	04:21
Air Blank		0.000	04:22
Diagnostics Check		OK	04:22

Data Download Date/Time: 02/26/2020 06:12

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAY,JOHN,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BELL,DEPUTY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: LEIDY

DATE/TIME OF TEST: 02/22/2020 20:00:32

Diagnostics Check		OK	19:53
Air Blank		0.000	19:53
Control Test		0.080	19:54
Air Blank		0.000	19:54
Subject Sample	# 1	0.183	19:55
Breath Volume		2.015	19:55
Air Blank		0.000	19:56
Air Blank		0.000	19:58
Subject Sample	# 2	0.174	19:58
Breath Volume		1.957	19:58
Air Blank		0.000	19:59
Control Test		0.078	19:59
Air Blank		0.000	20:00
Diagnostics Check		OK	20:00

Data Download Date/Time: 02/26/2020 06:12

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEPUTY,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SANDERS,DEPUTY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:16

SUBJECT'S LAST NAME: CARLOS-PUENTES

DATE/TIME OF TEST: 02/22/2020 21:42:36

Diagnostics Check		OK	21:36
Air Blank		0.000	21:36
Control Test		0.080	21:37
Air Blank		0.000	21:37
Subject Sample	# 1	0.147	21:38
Breath Volume		3.214	21:38
Air Blank		0.000	21:38
Air Blank		0.000	21:40
Subject Sample	# 2	0.146	21:41
Breath Volume		2.378	21:41
Air Blank		0.000	21:41
Control Test		0.079	21:41
Air Blank		0.000	21:42
Diagnostics Check		OK	21:42

Data Download Date/Time: 02/26/2020 06:12

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAMORS,JASON,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SANDERS,J
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: NOEL

DATE/TIME OF TEST: 02/23/2020 04:05:11

Diagnostics Check		OK	03:58
Air Blank		0.000	03:59
Control Test		0.080	03:59
Air Blank		0.000	03:59
Subject Sample	# 1	0.175	04:00
Breath Volume		2.523	04:00
Air Blank		0.000	04:01
Air Blank		0.000	04:02
Subject Sample	# 2	0.178	04:03
Breath Volume		2.214	04:03
Air Blank		0.000	04:04
Control Test		0.080	04:04
Air Blank		0.000	04:05
Diagnostics Check		OK	04:05

Data Download Date/Time: 02/26/2020 06:12

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ECKERT,GABRIEL,B
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: ECKERT,GABRIEL
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: FIGARSKY

DATE/TIME OF TEST: 02/25/2020 22:06:10

Diagnostics Check		OK	21:57
Air Blank		0.000	21:58
Control Test		0.079	21:58
Air Blank		0.000	21:58
Subject Sample	# 1	0.170	22:00
Breath Volume		2.171	22:00
Air Blank		0.000	22:01
Air Blank		0.000	22:03
Subject Sample	# 2	0.166	22:04
Breath Volume		1.636	22:04
Air Blank		0.000	22:05
Control Test		0.077	22:05
Air Blank		0.000	22:06
Diagnostics Check		OK	22:06

Data Download Date/Time: 02/26/2020 06:12

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAY,JOHN,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BELL,DEPUTY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:01

SUBJECT'S LAST NAME: HURLEY

DATE/TIME OF TEST: 02/27/2020 18:35:49

Diagnostics Check		OK	18:29
Air Blank		0.000	18:30
Control Test		0.079	18:30
Air Blank		0.000	18:31
Subject Sample	# 1	0.000	18:31
Breath Volume		3.375	18:31
Air Blank		0.000	18:32
Air Blank		0.000	18:34
Subject Sample	# 2	0.000	18:34
Breath Volume		2.339	18:34
Air Blank		0.000	18:34
Control Test		0.079	18:35
Air Blank		0.000	18:35
Diagnostics Check		OK	18:35

Data Download Date/Time: 03/26/2020 09:12

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAY,JOHN,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SANDERS,DEPUTY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: DYBER

DATE/TIME OF TEST: 02/28/2020 02:52:19

Diagnostics Check		OK	02:46
Air Blank		0.000	02:46
Control Test		0.081	02:46
Air Blank		0.000	02:47
Subject Sample	# 1	0.184	02:47
Breath Volume		1.640	02:47
Air Blank		0.000	02:48
Air Blank		0.000	02:50
Subject Sample	# 2	0.169	02:50
Breath Volume		1.398	02:50
Air Blank		0.000	02:51
Control Test		0.076	02:51
Air Blank		0.000	02:52
Diagnostics Check		OK	02:52

Data Download Date/Time: 03/26/2020 09:12

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDEZ,DEPUTY,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: ROBERTS,OFFICER
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 09:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:17

SUBJECT'S LAST NAME: LINDBERG

DATE/TIME OF TEST: 02/28/2020 09:43:50

Diagnostics Check		OK	09:37
Air Blank		0.000	09:37
Control Test		0.078	09:37
Air Blank		0.000	09:38
Subject Sample	# 1	0.179	09:39
Breath Volume		4.484	09:39
Air Blank		0.000	09:39
Air Blank		0.000	09:41
Subject Sample	# 2	0.181	09:42
Breath Volume		3.843	09:42
Air Blank		0.000	09:42
Control Test		0.078	09:43
Air Blank		0.000	09:43
Diagnostics Check		OK	09:43

Data Download Date/Time: 03/26/2020 09:12

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,N,N
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MONTANEZ,V
ARRESTING OFFICER AGENCY: Venice PD
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: SINNER

DATE/TIME OF TEST: 02/29/2020 03:22:19

Diagnostics Check		OK	03:15
Air Blank		0.000	03:16
Control Test		0.081	03:16
Air Blank		0.000	03:17
Subject Sample	# 1	0.054	03:17
Breath Volume		2.585	03:17
Air Blank		0.000	03:18
Air Blank		0.000	03:20
Subject Sample	# 2	0.058	03:20
Breath Volume		2.507	03:20
Air Blank		0.000	03:21
Control Test		0.080	03:21
Air Blank		0.000	03:22
Diagnostics Check		OK	03:22

Data Download Date/Time: 03/26/2020 09:12

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERRY,WAYNE,G
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: KOLLER,JOHN
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 18:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 02/06/2020 21:49:32

Diagnostics Check		OK	21:42
Air Blank		0.000	21:42
Control Test		0.080	21:42
Air Blank		0.000	21:43
Subject Sample	# 1	0.189	21:44
Breath Volume		1.371	21:44
Air Blank		0.000	21:45
Air Blank		0.000	21:46
Subject Sample	# 2	0.198	21:48
Breath Volume		2.539	21:48
Air Blank		0.000	21:48
Control Test		0.081	21:48
Air Blank		0.000	21:49
Diagnostics Check		OK	21:49

Data Download Date/Time: 02/26/2020 11:07

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERRY,WAYNE,G
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: CROWTHER,TREVOR
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: BASS

DATE/TIME OF TEST: 02/10/2020 23:29:09

Diagnostics Check		OK	23:16	
Air Blank		0.000	23:16	
Control Test		0.080	23:16	
Air Blank		0.000	23:17	
Subject Sample	# 1	0.010	23:20	Volume Not Met
Breath Volume		0.113	23:20	
Air Blank		0.000	23:21	
Air Blank		0.000	23:23	
Subject Sample	# 2	0.010	23:23	
Breath Volume		1.738	23:23	
Air Blank		0.000	23:24	
Air Blank		0.000	23:26	
Subject Sample	# 3	0.010	23:27	
Breath Volume		2.519	23:27	
Air Blank		0.000	23:28	
Control Test		0.082	23:28	
Air Blank		0.000	23:29	
Diagnostics Check		OK	23:29	

Data Download Date/Time: 02/26/2020 11:07

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERRY,WAYNE,G
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: BROWN,JOSH
ARRESTING OFFICER AGENCY: Umatilla PD
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 19:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: MCCOY

DATE/TIME OF TEST: 02/11/2020 20:21:09

Diagnostics Check		OK	20:06	
Air Blank		0.000	20:06	
Control Test		0.080	20:07	
Air Blank		0.000	20:07	
Subject Sample	# 1	0.173	20:10	Volume Not Met
Breath Volume		0.839	20:10	
Air Blank		0.000	20:11	
Air Blank		0.000	20:13	
Subject Sample	# 2	0.171	20:15	
Breath Volume		1.277	20:15	
Air Blank		0.000	20:16	
Air Blank		0.000	20:18	
Subject Sample	# 3	0.176	20:19	
Breath Volume		1.945	20:19	
Air Blank		0.000	20:20	
Control Test		0.079	20:20	
Air Blank		0.000	20:21	
Diagnostics Check		OK	20:21	

Data Download Date/Time: 02/26/2020 11:08

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,DAVID,A
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: PARKER,STEPHEN
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: MOHAMMED

DATE/TIME OF TEST: 02/22/2020 03:19:29

Diagnostics Check		OK	03:10
Air Blank		0.000	03:10
Control Test		0.082	03:10
Air Blank		0.000	03:11
Subject Sample	# 1	0.271	03:11
Breath Volume		2.523	03:11
Air Blank		0.000	03:12
Air Blank		0.000	03:14
Subject Sample	# 2	0.243	03:15
Breath Volume		2.132	03:15
Air Blank		0.000	03:15
Air Blank		0.000	03:17
Subject Sample	# 3	0.266	03:17
Breath Volume		2.812	03:17
Air Blank		0.000	03:18
Control Test		0.081	03:18
Air Blank		0.000	03:19
Diagnostics Check		OK	03:19

Data Download Date/Time: 02/26/2020 11:08

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,DAVID,A
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: GRIFFITH,MICHAEL
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 21:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:29

SUBJECT'S LAST NAME: HARTMAN

DATE/TIME OF TEST: 02/23/2020 21:58:43

Diagnostics Check		OK	21:55	
Air Blank		0.000	21:56	
Control Test		0.080	21:56	
Air Blank		0.000	21:57	
Subject Sample	# 1		21:57	Subject Test Refused
Breath Volume		0.000	21:57	
Air Blank		0.000	21:57	
Control Test		0.080	21:58	
Air Blank		0.000	21:58	
Diagnostics Check		OK	21:58	

Data Download Date/Time: 02/26/2020 11:08

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERRY,WAYNE,G
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: CROWTHER,TREVOR
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: CURRERI

DATE/TIME OF TEST: 02/25/2020 03:05:04

Diagnostics Check	OK	03:03	
Air Blank	0.000	03:03	
Control Test	0.081	03:04	
Air Blank		03:04	RFI Detect
Air Blank	0.000	03:05	

Data Download Date/Time: 02/26/2020 11:08

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERRY,WAYNE,G
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: CROWTHER,TREVOR
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: CURRERI

DATE/TIME OF TEST: 02/25/2020 03:18:02

Diagnostics Check		OK	03:10
Air Blank		0.000	03:10
Control Test		0.081	03:10
Air Blank		0.000	03:11
Subject Sample	# 1	0.163	03:12
Breath Volume		1.402	03:12
Air Blank		0.000	03:13
Air Blank		0.000	03:15
Subject Sample	# 2	0.175	03:16
Breath Volume		1.414	03:16
Air Blank		0.000	03:17
Control Test		0.080	03:17
Air Blank		0.000	03:17
Diagnostics Check		OK	03:18

Data Download Date/Time: 02/26/2020 11:08

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DEP,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MILLER,DEP R
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: PEJNOVIC

DATE/TIME OF TEST: 02/04/2020 23:29:00

Diagnostics Check		OK	23:23
Air Blank		0.000	23:23
Control Test		0.079	23:24
Air Blank		0.000	23:24
Subject Sample	# 1	0.000	23:24
Breath Volume		3.625	23:24
Air Blank		0.000	23:25
Air Blank		0.000	23:27
Subject Sample	# 2	0.000	23:27
Breath Volume		3.937	23:27
Air Blank		0.000	23:28
Control Test		0.079	23:28
Air Blank		0.000	23:28
Diagnostics Check		OK	23:29

Data Download Date/Time: 02/26/2020 06:20

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DEP,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BELL,DEP D
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: WOLTER

DATE/TIME OF TEST: 02/08/2020 20:16:28

Diagnostics Check		OK	20:10
Air Blank		0.000	20:10
Control Test		0.079	20:11
Air Blank		0.000	20:11
Subject Sample	# 1	0.176	20:12
Breath Volume		2.867	20:12
Air Blank		0.000	20:12
Air Blank		0.000	20:14
Subject Sample	# 2	0.166	20:14
Breath Volume		2.890	20:14
Air Blank		0.000	20:15
Control Test		0.080	20:15
Air Blank		0.000	20:16
Diagnostics Check		OK	20:16

Data Download Date/Time: 02/26/2020 06:20

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DEP,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MILLER,DEP R
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:39

SUBJECT'S LAST NAME: JARVIS

DATE/TIME OF TEST: 02/08/2020 22:05:45

Diagnostics Check		OK	21:59
Air Blank		0.000	22:00
Control Test		0.080	22:00
Air Blank		0.000	22:01
Subject Sample	# 1	0.000	22:01
Breath Volume		2.867	22:01
Air Blank		0.000	22:02
Air Blank		0.000	22:04
Subject Sample	# 2	0.000	22:04
Breath Volume		2.882	22:04
Air Blank		0.000	22:04
Control Test		0.080	22:05
Air Blank		0.000	22:05
Diagnostics Check		OK	22:05

Data Download Date/Time: 02/26/2020 06:20

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DEP,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BELL,DEP D
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 02/09/2020 02:06:31

Diagnostics Check		OK	02:00
Air Blank		0.000	02:00
Control Test		0.079	02:01
Air Blank		0.000	02:01
Subject Sample	# 1	0.166	02:02
Breath Volume		3.000	02:02
Air Blank		0.000	02:02
Air Blank		0.000	02:04
Subject Sample	# 2	0.168	02:05
Breath Volume		2.707	02:05
Air Blank		0.000	02:05
Control Test		0.080	02:05
Air Blank		0.000	02:06
Diagnostics Check		OK	02:06

Data Download Date/Time: 02/26/2020 06:20

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DEP,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MILLER,DEP R
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: LADAK-BARBER

DATE/TIME OF TEST: 02/09/2020 03:47:33

Diagnostics Check		OK	03:39
Air Blank		0.000	03:40
Control Test		0.080	03:40
Air Blank		0.000	03:41
Subject Sample	# 1	0.135	03:42
Breath Volume		1.714	03:42
Air Blank		0.000	03:43
Air Blank		0.000	03:45
Subject Sample	# 2	0.129	03:46
Breath Volume		1.593	03:46
Air Blank		0.000	03:46
Control Test		0.079	03:46
Air Blank		0.000	03:47
Diagnostics Check		OK	03:47

Data Download Date/Time: 02/26/2020 06:20

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DEP,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MILLER,DEP R
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: MARRERO

DATE/TIME OF TEST: 02/12/2020 19:07:23

Diagnostics Check		OK	19:00
Air Blank		0.000	19:01
Control Test		0.079	19:01
Air Blank		0.000	19:02
Subject Sample	# 1	0.203	19:02
Breath Volume		2.328	19:02
Air Blank		0.000	19:03
Air Blank		0.000	19:04
Subject Sample	# 2	0.185	19:05
Breath Volume		2.335	19:05
Air Blank		0.000	19:06
Control Test		0.078	19:06
Air Blank		0.000	19:07
Diagnostics Check		OK	19:07

Data Download Date/Time: 02/26/2020 06:20

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DEP,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MILLER,DEP R
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: BERG

DATE/TIME OF TEST: 02/14/2020 01:40:52

Diagnostics Check		OK	01:34
Air Blank		0.000	01:34
Control Test		0.079	01:35
Air Blank		0.000	01:35
Subject Sample	# 1	0.122	01:36
Breath Volume		3.453	01:36
Air Blank		0.000	01:37
Air Blank		0.000	01:38
Subject Sample	# 2	0.120	01:39
Breath Volume		2.031	01:39
Air Blank		0.000	01:39
Control Test		0.078	01:40
Air Blank		0.000	01:40
Diagnostics Check		OK	01:40

Data Download Date/Time: 02/26/2020 06:20

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DEP,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MILLER,DEP R
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: ROSENBAUM

DATE/TIME OF TEST: 02/17/2020 20:45:20

Diagnostics Check		OK	20:39
Air Blank		0.000	20:39
Control Test		0.078	20:40
Air Blank		0.000	20:40
Subject Sample	# 1	0.086	20:41
Breath Volume		1.476	20:41
Air Blank		0.000	20:41
Air Blank		0.000	20:43
Subject Sample	# 2	0.085	20:43
Breath Volume		2.664	20:43
Air Blank		0.000	20:44
Control Test		0.078	20:44
Air Blank		0.000	20:45
Diagnostics Check		OK	20:45

Data Download Date/Time: 02/26/2020 06:20

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,J,S
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BARR,J
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: RIENDEAU

DATE/TIME OF TEST: 02/02/2020 23:35:59

Diagnostics Check		OK	23:28
Air Blank		0.000	23:29
Control Test		0.079	23:29
Air Blank		0.000	23:30
Subject Sample	# 1	0.309	23:31
Breath Volume		1.781	23:31
Air Blank		0.000	23:32
Air Blank		0.000	23:33
Subject Sample	# 2	0.292	23:34
Breath Volume		2.101	23:34
Air Blank		0.000	23:35
Control Test		0.080	23:35
Air Blank		0.000	23:35
Diagnostics Check		OK	23:35

Data Download Date/Time: 02/26/2020 06:16

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,JEREMY,S
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BARR,J
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:21

SUBJECT'S LAST NAME: BLAISDELL

DATE/TIME OF TEST: 02/16/2020 20:55:45

Diagnostics Check		OK	20:49
Air Blank		0.000	20:50
Control Test		0.079	20:50
Air Blank		0.000	20:50
Subject Sample	# 1	0.204	20:51
Breath Volume		2.210	20:51
Air Blank		0.000	20:51
Air Blank		0.000	20:53
Subject Sample	# 2	0.205	20:54
Breath Volume		2.265	20:54
Air Blank		0.000	20:54
Control Test		0.079	20:55
Air Blank		0.000	20:55
Diagnostics Check		OK	20:55

Data Download Date/Time: 02/26/2020 06:16

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,J,S
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BARR,J
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 02/19/2020 19:32:14

Diagnostics Check		OK	19:26
Air Blank		0.000	19:26
Control Test		0.079	19:26
Air Blank		0.000	19:27
Subject Sample	# 1	0.145	19:27
Breath Volume		2.113	19:27
Air Blank		0.000	19:28
Air Blank		0.000	19:30
Subject Sample	# 2	0.145	19:30
Breath Volume		2.621	19:30
Air Blank		0.000	19:31
Control Test		0.078	19:31
Air Blank		0.000	19:32
Diagnostics Check		OK	19:32

Data Download Date/Time: 02/26/2020 06:16

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAIR,TIMOTHY,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: BLAIR,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 16:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:38

SUBJECT'S LAST NAME: DEYOUNG

DATE/TIME OF TEST: 02/12/2020 17:05:11

Diagnostics Check		OK	16:58
Air Blank		0.000	16:58
Control Test		0.081	16:58
Air Blank		0.000	16:59
Subject Sample	# 1	0.000	17:00
Breath Volume		2.054	17:00
Air Blank		0.000	17:01
Air Blank		0.000	17:03
Subject Sample	# 2	0.000	17:03
Breath Volume		1.234	17:03
Air Blank		0.000	17:04
Control Test		0.080	17:04
Air Blank		0.000	17:05
Diagnostics Check		OK	17:05

Data Download Date/Time: 02/29/2020 22:09

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLFF,ROSE,M
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: DULIN,HERBERT
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 11:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:22

SUBJECT'S LAST NAME: TODD

DATE/TIME OF TEST: 02/13/2020 12:54:35

Diagnostics Check		OK	12:48
Air Blank		0.000	12:49
Control Test		0.081	12:49
Air Blank		0.000	12:49
Subject Sample	# 1	0.000	12:50
Breath Volume		2.457	12:50
Air Blank		0.000	12:50
Air Blank		0.000	12:52
Subject Sample	# 2	0.000	12:53
Breath Volume		3.050	12:53
Air Blank		0.000	12:53
Control Test		0.081	12:54
Air Blank		0.000	12:54
Diagnostics Check		OK	12:54

Data Download Date/Time: 02/29/2020 22:09

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDEZ,JAMIE,J
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: MENDEZ,JAMIE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: HEWITT

DATE/TIME OF TEST: 02/19/2020 05:21:46

Diagnostics Check		OK	05:15
Air Blank		0.000	05:15
Control Test		0.081	05:16
Air Blank		0.000	05:16
Subject Sample	# 1	0.185	05:17
Breath Volume		2.398	05:17
Air Blank		0.000	05:17
Air Blank		0.000	05:19
Subject Sample	# 2	0.178	05:20
Breath Volume		1.523	05:20
Air Blank		0.000	05:20
Control Test		0.081	05:21
Air Blank		0.000	05:21
Diagnostics Check		OK	05:21

Data Download Date/Time: 02/29/2020 22:09

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAIR,TIMOTHY,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: BLAIR,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 02/23/2020 04:08:10

Diagnostics Check		OK	04:01
Air Blank		0.000	04:02
Control Test		0.081	04:02
Air Blank		0.000	04:03
Subject Sample	# 1	0.151	04:03
Breath Volume		2.582	04:03
Air Blank		0.000	04:04
Air Blank		0.000	04:06
Subject Sample	# 2	0.150	04:06
Breath Volume		1.527	04:06
Air Blank		0.000	04:07
Control Test		0.081	04:07
Air Blank		0.000	04:08
Diagnostics Check		OK	04:08

Data Download Date/Time: 02/29/2020 22:09

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAIR,TIMOTHY,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: BLAIR,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: WORKMAN

DATE/TIME OF TEST: 02/27/2020 04:49:14

Diagnostics Check		OK	04:42
Air Blank		0.000	04:42
Control Test		0.080	04:42
Air Blank		0.000	04:43
Subject Sample	# 1	0.000	04:44
Breath Volume		1.410	04:44
Air Blank		0.000	04:44
Air Blank		0.000	04:46
Subject Sample	# 2	0.000	04:47
Breath Volume		1.500	04:47
Air Blank		0.000	04:48
Control Test		0.081	04:48
Air Blank		0.000	04:49
Diagnostics Check		OK	04:49

Data Download Date/Time: 02/29/2020 22:09

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLFF,ROSE,M
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: WOLFF,ROSE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 13:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:05

SUBJECT'S LAST NAME: VAMMEN

DATE/TIME OF TEST: 02/29/2020 14:36:11

Diagnostics Check		OK	14:29
Air Blank		0.000	14:30
Control Test		0.082	14:30
Air Blank		0.000	14:31
Subject Sample	# 1	0.190	14:31
Breath Volume		2.789	14:31
Air Blank		0.000	14:32
Air Blank		0.000	14:34
Subject Sample	# 2	0.176	14:34
Breath Volume		1.902	14:34
Air Blank		0.000	14:35
Control Test		0.082	14:35
Air Blank		0.000	14:36
Diagnostics Check		OK	14:36

Data Download Date/Time: 02/29/2020 22:09

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEIMIG,EMORIE,I
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: PEREZ,GEORGE
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: HARRISON III

DATE/TIME OF TEST: 02/02/2020 04:44:54

Diagnostics Check		OK	04:33	
Air Blank		0.000	04:33	
Control Test		0.080	04:34	
Air Blank		0.000	04:34	
Subject Sample	# 1	0.142	04:37	Volume Not Met
Breath Volume		0.109	04:37	
Air Blank		0.000	04:38	
Air Blank		0.000	04:40	
Subject Sample	# 2	0.120	04:43	Volume Not Met
Breath Volume		0.234	04:43	
Air Blank		0.000	04:43	
Control Test		0.079	04:44	
Air Blank		0.000	04:44	
Diagnostics Check		OK	04:44	

Data Download Date/Time: 02/26/2020 01:49

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: FUENTES,MELISSA
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:54

SUBJECT'S LAST NAME: VOIGT

DATE/TIME OF TEST: 02/02/2020 20:30:08

Diagnostics Check		OK	20:16	
Air Blank		0.000	20:16	
Control Test		0.080	20:16	
Air Blank		0.000	20:17	
Subject Sample	# 1		20:20	No Sample Provided
Breath Volume		0.000	20:20	
Air Blank		0.000	20:20	
Air Blank		0.000	20:22	
Subject Sample	# 2	0.181	20:25	
Breath Volume		1.171	20:25	
Air Blank		0.000	20:25	
Air Blank		0.000	20:27	
Subject Sample	# 3	0.172	20:28	
Breath Volume		1.300	20:28	
Air Blank		0.000	20:29	
Control Test		0.079	20:29	
Air Blank		0.000	20:30	
Diagnostics Check		OK	20:30	

Data Download Date/Time: 02/26/2020 01:49

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEIMIG,EMORIE,I
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: JOHNSON,TRAVIS
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: MCCRAY

DATE/TIME OF TEST: 02/05/2020 21:53:20

Diagnostics Check		OK	21:42
Air Blank		0.000	21:43
Control Test		0.080	21:43
Air Blank		0.000	21:44
Subject Sample	# 1	0.192	21:45
Breath Volume		1.753	21:45
Air Blank		0.000	21:45
Air Blank		0.000	21:47
Subject Sample	# 2	0.169	21:48
Breath Volume		1.742	21:48
Air Blank		0.000	21:49
Air Blank		0.000	21:50
Subject Sample	# 3	0.177	21:51
Breath Volume		2.101	21:51
Air Blank		0.000	21:52
Control Test		0.079	21:52
Air Blank		0.000	21:53
Diagnostics Check		OK	21:53

Data Download Date/Time: 02/26/2020 01:50

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELCH,CHAD,J
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: SHRIGLEY,ROBERT
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: STUCKEY

DATE/TIME OF TEST: 02/09/2020 03:20:28

Diagnostics Check		OK	03:14
Air Blank		0.000	03:14
Control Test		0.081	03:15
Air Blank		0.000	03:15
Subject Sample	# 1	0.000	03:16
Breath Volume		2.128	03:16
Air Blank		0.000	03:16
Air Blank		0.000	03:18
Subject Sample	# 2	0.000	03:19
Breath Volume		2.171	03:19
Air Blank		0.000	03:19
Control Test		0.080	03:19
Air Blank		0.000	03:20
Diagnostics Check		OK	03:20

Data Download Date/Time: 02/26/2020 01:50

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: EDWIN

DATE/TIME OF TEST: 02/15/2020 04:07:46

Diagnostics Check		OK	04:01
Air Blank		0.000	04:02
Control Test		0.080	04:02
Air Blank		0.000	04:02
Subject Sample	# 1	0.222	04:03
Breath Volume		2.023	04:03
Air Blank		0.000	04:03
Air Blank		0.000	04:05
Subject Sample	# 2	0.221	04:06
Breath Volume		2.000	04:06
Air Blank		0.000	04:06
Control Test		0.081	04:07
Air Blank		0.000	04:07
Diagnostics Check		OK	04:07

Data Download Date/Time: 02/26/2020 01:50

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUYTON,JOHN,A
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GUYTON,JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:03

SUBJECT'S LAST NAME: GRIMMAGE

DATE/TIME OF TEST: 02/17/2020 22:35:20

Diagnostics Check		OK	22:27
Air Blank		0.000	22:28
Control Test		0.080	22:28
Air Blank		0.000	22:29
Subject Sample	# 1	0.202	22:30
Breath Volume		1.675	22:30
Air Blank		0.000	22:31
Air Blank		0.000	22:32
Subject Sample	# 2	0.201	22:33
Breath Volume		1.437	22:33
Air Blank		0.000	22:34
Control Test		0.079	22:34
Air Blank		0.000	22:35
Diagnostics Check		OK	22:35

Data Download Date/Time: 02/26/2020 01:50

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: CONNOR

DATE/TIME OF TEST: 02/21/2020 04:27:40

Diagnostics Check		OK	04:21
Air Blank		0.000	04:22
Control Test		0.080	04:22
Air Blank		0.000	04:22
Subject Sample	# 1	0.159	04:23
Breath Volume		1.679	04:23
Air Blank		0.000	04:23
Air Blank		0.000	04:25
Subject Sample	# 2	0.156	04:26
Breath Volume		1.804	04:26
Air Blank		0.000	04:26
Control Test		0.080	04:27
Air Blank		0.000	04:27
Diagnostics Check		OK	04:27

Data Download Date/Time: 02/26/2020 01:50

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELCH,CHAD,J
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: SHRIGLEY,ROBERT
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: BURNEY

DATE/TIME OF TEST: 02/22/2020 04:35:51

Diagnostics Check		OK	04:29
Air Blank		0.000	04:30
Control Test		0.081	04:30
Air Blank		0.000	04:30
Subject Sample	# 1	0.194	04:31
Breath Volume		1.984	04:31
Air Blank		0.000	04:32
Air Blank		0.000	04:33
Subject Sample	# 2	0.190	04:34
Breath Volume		1.550	04:34
Air Blank		0.000	04:34
Control Test		0.079	04:35
Air Blank		0.000	04:35
Diagnostics Check		OK	04:35

Data Download Date/Time: 02/26/2020 01:50

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEIMIG,EMORIE,I
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: MILMAN,ANDREW
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: TURK

DATE/TIME OF TEST: 02/28/2020 23:43:08

Diagnostics Check		OK	23:36
Air Blank		0.000	23:37
Control Test		0.080	23:37
Air Blank		0.000	23:37
Subject Sample	# 1	0.183	23:38
Breath Volume		4.285	23:38
Air Blank		0.000	23:39
Air Blank		0.000	23:40
Subject Sample	# 2	0.180	23:41
Breath Volume		4.324	23:41
Air Blank		0.000	23:42
Control Test		0.080	23:42
Air Blank		0.000	23:43
Diagnostics Check		OK	23:43

Data Download Date/Time: 03/30/2020 03:10

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

INSTRUMENT REGISTERED TO: NAS MAYPORT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GLOSKA,EDWARD,J
BREATH TEST OPERATOR AGENCY: Mayport Naval Base
ARRESTING OFFICER: ALLEN,JOHNNY
ARRESTING OFFICER AGENCY: Mayport Naval Base
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: SAFFO-MILLER

DATE/TIME OF TEST: 02/03/2020 03:59:04

Diagnostics Check		OK	03:52
Air Blank		0.000	03:53
Control Test		0.080	03:53
Air Blank		0.000	03:54
Subject Sample	# 1	0.185	03:54
Breath Volume		1.707	03:54
Air Blank		0.000	03:55
Air Blank		0.000	03:57
Subject Sample	# 2	0.187	03:57
Breath Volume		2.687	03:57
Air Blank		0.000	03:58
Control Test		0.081	03:58
Air Blank		0.000	03:58
Diagnostics Check		OK	03:59

Data Download Date/Time: 02/11/2020 13:22

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

INSTRUMENT REGISTERED TO: NAS MAYPORT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OCASIO,JOSE,A
BREATH TEST OPERATOR AGENCY: Mayport Naval Base
ARRESTING OFFICER: STAATS,PHILLIP
ARRESTING OFFICER AGENCY: Mayport Naval Base
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: FLORE

DATE/TIME OF TEST: 02/07/2020 19:44:43

Diagnostics Check		OK	19:37
Air Blank		0.000	19:38
Control Test		0.078	19:38
Air Blank		0.000	19:38
Subject Sample	# 1	0.152	19:40
Breath Volume		3.398	19:40
Air Blank		0.000	19:40
Air Blank		0.000	19:42
Subject Sample	# 2	0.132	19:43
Breath Volume		1.890	19:43
Air Blank		0.000	19:43
Control Test		0.079	19:44
Air Blank		0.000	19:44
Diagnostics Check		OK	19:44

Data Download Date/Time: 02/11/2020 13:22

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

INSTRUMENT REGISTERED TO: NAS MAYPORT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,DEREK,V
BREATH TEST OPERATOR AGENCY: Mayport Naval Base
ARRESTING OFFICER: POLLOCK,DEREK
ARRESTING OFFICER AGENCY: Mayport Naval Base
LAST AGENCY INSPECTION: 02/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: REICHERT-CONTE

DATE/TIME OF TEST: 02/16/2020 03:48:40

Diagnostics Check		OK	03:41
Air Blank		0.000	03:41
Control Test		0.080	03:42
Air Blank		0.000	03:42
Subject Sample	# 1	0.192	03:44
Breath Volume		2.187	03:44
Air Blank		0.000	03:44
Air Blank		0.000	03:46
Subject Sample	# 2	0.193	03:47
Breath Volume		2.132	03:47
Air Blank		0.000	03:47
Control Test		0.079	03:48
Air Blank		0.000	03:48
Diagnostics Check		OK	03:48

Data Download Date/Time: 03/30/2020 14:08

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

INSTRUMENT REGISTERED TO: NAS MAYPORT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OCASIO,JOSE,A
BREATH TEST OPERATOR AGENCY: Mayport Naval Base
ARRESTING OFFICER: JILES,DARRYL
ARRESTING OFFICER AGENCY: Mayport Naval Base
LAST AGENCY INSPECTION: 02/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: THOMAS JR

DATE/TIME OF TEST: 02/23/2020 03:10:59

Diagnostics Check		OK	03:04
Air Blank		0.000	03:04
Control Test		0.079	03:04
Air Blank		0.000	03:05
Subject Sample	# 1	0.249	03:06
Breath Volume		2.417	03:06
Air Blank		0.000	03:06
Air Blank		0.000	03:08
Subject Sample	# 2	0.240	03:09
Breath Volume		2.851	03:09
Air Blank		0.000	03:10
Control Test		0.079	03:10
Air Blank		0.000	03:10
Diagnostics Check		OK	03:10

Data Download Date/Time: 03/30/2020 14:08

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KNIGHT,RYAN
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 02/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: BELLORA

DATE/TIME OF TEST: 02/12/2020 03:38:48

Diagnostics Check		OK	03:23	
Air Blank		0.000	03:24	
Control Test		0.082	03:24	
Air Blank		0.000	03:25	
Subject Sample	# 1	0.138	03:28	Volume Not Met
Breath Volume		0.140	03:28	
Air Blank		0.000	03:29	
Air Blank		0.000	03:30	
Subject Sample	# 2	0.138	03:31	
Breath Volume		1.285	03:31	
Air Blank		0.000	03:31	
Air Blank		0.000	03:33	
Subject Sample	# 3	0.143	03:36	
Breath Volume		1.203	03:36	
Air Blank		0.000	03:37	
Control Test		0.082	03:38	
Air Blank		0.000	03:38	
Diagnostics Check		OK	03:38	

Data Download Date/Time: 02/20/2020 16:57

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SANCHEZ,JOSE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 02/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: CRENSHAW

DATE/TIME OF TEST: 02/15/2020 02:28:54

Diagnostics Check		OK	02:22
Air Blank		0.000	02:22
Control Test		0.082	02:22
Air Blank		0.000	02:23
Subject Sample	# 1	0.247	02:24
Breath Volume		2.625	02:24
Air Blank		0.000	02:24
Air Blank		0.000	02:26
Subject Sample	# 2	0.255	02:27
Breath Volume		2.859	02:27
Air Blank		0.000	02:27
Control Test		0.081	02:28
Air Blank		0.000	02:28
Diagnostics Check		OK	02:28

Data Download Date/Time: 02/20/2020 16:57

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RAMEY,MICHAEL
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: GOLDSMITH

DATE/TIME OF TEST: 02/16/2020 04:02:38

Diagnostics Check		OK	03:58	
Air Blank		0.000	03:59	
Control Test		0.081	03:59	
Air Blank		0.000	04:00	
Subject Sample	# 1		04:01	Subject Test Refused
Breath Volume		0.000	04:01	
Air Blank		0.000	04:01	
Control Test		0.082	04:02	
Air Blank		0.000	04:02	
Diagnostics Check		OK	04:02	

Data Download Date/Time: 02/20/2020 16:57

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HALL,JONATHAN
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 02/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: MENDOZA-RAMIREZ

DATE/TIME OF TEST: 02/17/2020 05:32:53

Diagnostics Check		OK	05:25
Air Blank		0.000	05:26
Control Test		0.081	05:26
Air Blank		0.000	05:27
Subject Sample	# 1	0.265	05:27
Breath Volume		2.238	05:27
Air Blank		0.000	05:28
Air Blank		0.000	05:30
Subject Sample	# 2	0.263	05:31
Breath Volume		1.378	05:31
Air Blank		0.000	05:31
Control Test		0.081	05:32
Air Blank		0.000	05:32
Diagnostics Check		OK	05:32

Data Download Date/Time: 02/20/2020 16:57

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MARTINEZ,ERNESTO
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: BAILEY-WILLIAMS

DATE/TIME OF TEST: 02/19/2020 20:33:48

Diagnostics Check		OK	20:27
Air Blank		0.000	20:28
Control Test		0.082	20:28
Air Blank		0.000	20:28
Subject Sample	# 1	0.000	20:29
Breath Volume		3.035	20:29
Air Blank		0.000	20:29
Air Blank		0.000	20:31
Subject Sample	# 2	0.000	20:32
Breath Volume		3.410	20:32
Air Blank		0.000	20:32
Control Test		0.080	20:33
Air Blank		0.000	20:33
Diagnostics Check		OK	20:33

Data Download Date/Time: 02/20/2020 16:57

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SOUD,AHMAD
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 02/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: DURDEN

DATE/TIME OF TEST: 02/22/2020 04:34:01

Diagnostics Check		OK	04:22	
Air Blank		0.000	04:23	
Control Test		0.082	04:23	
Air Blank		0.000	04:23	
Subject Sample	# 1		04:27	No Sample Provided
Breath Volume		0.000	04:27	
Air Blank		0.000	04:27	
Air Blank		0.000	04:29	
Subject Sample	# 2		04:32	No Sample Provided
Breath Volume		0.000	04:32	
Air Blank		0.000	04:33	
Control Test		0.083	04:33	
Air Blank		0.000	04:33	
Diagnostics Check		OK	04:34	

Data Download Date/Time: 02/27/2020 09:09

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAWKINS,DEBBIE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: RUISE

DATE/TIME OF TEST: 02/23/2020 00:29:42

Diagnostics Check		OK	00:19
Air Blank		0.000	00:19
Control Test		0.082	00:19
Air Blank		0.000	00:20
Subject Sample	# 1	0.163	00:22
Breath Volume		1.457	00:22
Air Blank		0.000	00:23
Air Blank		0.000	00:24
Subject Sample	# 2	0.158	00:28
Breath Volume		2.367	00:28
Air Blank		0.000	00:28
Control Test		0.080	00:29
Air Blank		0.000	00:29
Diagnostics Check		OK	00:29

Data Download Date/Time: 02/27/2020 09:09

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAYS,GARION
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 02/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: ULICNY

DATE/TIME OF TEST: 02/27/2020 01:46:07

Diagnostics Check		OK	01:39
Air Blank		0.000	01:39
Control Test		0.082	01:40
Air Blank		0.000	01:40
Subject Sample	# 1	0.204	01:41
Breath Volume		2.289	01:41
Air Blank		0.000	01:41
Air Blank		0.000	01:43
Subject Sample	# 2	0.201	01:44
Breath Volume		2.097	01:44
Air Blank		0.000	01:45
Control Test		0.082	01:45
Air Blank		0.000	01:46
Diagnostics Check		OK	01:46

Data Download Date/Time: 02/27/2020 09:09

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN',WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KELEMEN,JANOS
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: MURILLO

DATE/TIME OF TEST: 02/29/2020 01:45:12

Diagnostics Check		OK	01:39
Air Blank		0.000	01:39
Control Test		0.082	01:39
Air Blank		0.000	01:40
Subject Sample	# 1	0.174	01:40
Breath Volume		2.558	01:40
Air Blank		0.000	01:41
Air Blank		0.000	01:43
Subject Sample	# 2	0.170	01:43
Breath Volume		2.750	01:43
Air Blank		0.000	01:44
Control Test		0.081	01:44
Air Blank		0.000	01:45
Diagnostics Check		OK	01:45

Data Download Date/Time: 03/05/2020 08:29

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ROJAS,YOAN
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 02/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: COTO

DATE/TIME OF TEST: 02/29/2020 03:31:37

Diagnostics Check		OK	03:25
Air Blank		0.000	03:25
Control Test		0.082	03:25
Air Blank		0.000	03:26
Subject Sample	# 1	0.137	03:26
Breath Volume		2.714	03:26
Air Blank		0.000	03:27
Air Blank		0.000	03:29
Subject Sample	# 2	0.133	03:30
Breath Volume		2.730	03:30
Air Blank		0.000	03:30
Control Test		0.081	03:31
Air Blank		0.000	03:31
Diagnostics Check		OK	03:31

Data Download Date/Time: 03/05/2020 08:29

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 02/14/2020 22:43:20

Diagnostics Check		OK	22:34	
Air Blank		0.000	22:35	
Control Test		0.079	22:35	
Air Blank		0.000	22:36	
Subject Sample	# 1		22:39	No Sample Provided
Breath Volume		0.000	22:39	
Air Blank		0.000	22:39	
Air Blank		0.000	22:41	
Subject Sample	# 2		22:41	Subject Test Refused
Breath Volume		0.000	22:41	
Air Blank		0.000	22:42	
Control Test		0.077	22:42	
Air Blank		0.000	22:43	
Diagnostics Check		OK	22:43	

Data Download Date/Time: 03/06/2020 15:29

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: MAHONEY

DATE/TIME OF TEST: 02/15/2020 19:42:11

Diagnostics Check		OK	19:35
Air Blank		0.000	19:36
Control Test		0.078	19:36
Air Blank		0.000	19:37
Subject Sample	# 1	0.167	19:37
Breath Volume		1.496	19:37
Air Blank		0.000	19:38
Air Blank		0.000	19:40
Subject Sample	# 2	0.173	19:40
Breath Volume		1.269	19:40
Air Blank		0.000	19:41
Control Test		0.077	19:41
Air Blank		0.000	19:42
Diagnostics Check		OK	19:42

Data Download Date/Time: 03/06/2020 15:29

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: GLENN

DATE/TIME OF TEST: 02/16/2020 02:20:29

Diagnostics Check		OK	02:14
Air Blank		0.000	02:14
Control Test		0.079	02:15
Air Blank		0.000	02:15
Subject Sample	# 1	0.209	02:16
Breath Volume		2.976	02:16
Air Blank		0.000	02:16
Air Blank		0.000	02:18
Subject Sample	# 2	0.206	02:18
Breath Volume		2.617	02:18
Air Blank		0.000	02:19
Control Test		0.078	02:19
Air Blank		0.000	02:20
Diagnostics Check		OK	02:20

Data Download Date/Time: 03/06/2020 15:29

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: BENSON

DATE/TIME OF TEST: 02/17/2020 00:35:22

Diagnostics Check		OK	00:27
Air Blank		0.000	00:28
Control Test		0.076	00:28
Air Blank		0.000	00:29
Subject Sample	# 1	0.190	00:30
Breath Volume		2.375	00:30
Air Blank		0.000	00:30
Air Blank		0.000	00:32
Subject Sample	# 2	0.181	00:33
Breath Volume		2.277	00:33
Air Blank		0.000	00:34
Control Test		0.078	00:34
Air Blank		0.000	00:35
Diagnostics Check		OK	00:35

Data Download Date/Time: 03/06/2020 15:29

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVAS,W,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: SILL,N
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 01/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: MORENO GONZALEZ

DATE/TIME OF TEST: 02/01/2020 00:05:41

Diagnostics Check		OK	23:59
Air Blank		0.000	00:00
Control Test		0.080	00:00
Air Blank		0.000	00:00
Subject Sample	# 1	0.087	00:01
Breath Volume		2.265	00:01
Air Blank		0.000	00:01
Air Blank		0.000	00:03
Subject Sample	# 2	0.085	00:04
Breath Volume		1.652	00:04
Air Blank		0.000	00:04
Control Test		0.080	00:05
Air Blank		0.000	00:05
Diagnostics Check		OK	00:05

Data Download Date/Time: 02/03/2020 16:18

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEORGE,SHANNON,N
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: WALBERT,A
ARRESTING OFFICER AGENCY: NAPLES PD
LAST AGENCY INSPECTION: 01/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: MIHALOVICH

DATE/TIME OF TEST: 02/01/2020 01:32:09

Diagnostics Check		OK	01:24
Air Blank		0.000	01:25
Control Test		0.079	01:25
Air Blank		0.000	01:25
Subject Sample	# 1	0.257	01:26
Breath Volume		2.898	01:26
Air Blank		0.000	01:27
Air Blank		0.000	01:29
Subject Sample	# 2	0.255	01:30
Breath Volume		2.332	01:30
Air Blank		0.000	01:31
Control Test		0.079	01:31
Air Blank		0.000	01:32
Diagnostics Check		OK	01:32

Data Download Date/Time: 02/03/2020 16:18

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLANEDA ROMO,KARLA,A
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: SALLS,LETTY
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 01/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:32

SUBJECT'S LAST NAME: QUINTERO DIAZ

DATE/TIME OF TEST: 02/01/2020 21:07:31

Diagnostics Check		OK	20:54
Air Blank		0.000	20:55
Control Test		0.079	20:55
Air Blank		0.000	20:56
Subject Sample	# 1	0.397	20:56
Breath Volume		1.617	20:56
Air Blank		0.000	20:57
Air Blank		0.000	20:59
Subject Sample	# 2	0.374	21:01
Breath Volume		1.281	21:01
Air Blank		0.000	21:02
Air Blank		0.000	21:04
Subject Sample	# 3	0.355	21:05
Breath Volume		1.375	21:05
Air Blank		0.000	21:06
Control Test		0.076	21:06
Air Blank		0.000	21:07
Diagnostics Check		OK	21:07

Data Download Date/Time: 02/03/2020 16:18

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUGO,JONATHAN,B
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: NOVA,M
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 01/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: JEFFERIES

DATE/TIME OF TEST: 02/02/2020 04:01:49

Diagnostics Check		OK	03:55
Air Blank		0.000	03:56
Control Test		0.080	03:56
Air Blank		0.000	03:56
Subject Sample	# 1	0.201	03:57
Breath Volume		2.457	03:57
Air Blank		0.000	03:58
Air Blank		0.000	03:59
Subject Sample	# 2	0.197	04:00
Breath Volume		2.425	04:00
Air Blank		0.000	04:00
Control Test		0.079	04:01
Air Blank		0.000	04:01
Diagnostics Check		OK	04:01

Data Download Date/Time: 02/03/2020 16:18

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODI,DOMENIC,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: ROWLEY,C
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 01/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: DOBYANSKI

DATE/TIME OF TEST: 02/03/2020 02:27:58

Diagnostics Check		OK	02:21
Air Blank		0.000	02:21
Control Test		0.080	02:22
Air Blank		0.000	02:22
Subject Sample	# 1	0.209	02:23
Breath Volume		1.417	02:23
Air Blank		0.000	02:24
Air Blank		0.000	02:25
Subject Sample	# 2	0.193	02:26
Breath Volume		1.535	02:26
Air Blank		0.000	02:27
Control Test		0.079	02:27
Air Blank		0.000	02:27
Diagnostics Check		OK	02:27

Data Download Date/Time: 02/03/2020 16:18

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOEL,Y,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: KAZMIERCZAK,B
ARRESTING OFFICER AGENCY: NAPLES PD
LAST AGENCY INSPECTION: 01/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: PETROVSKA

DATE/TIME OF TEST: 02/03/2020 03:52:51

Diagnostics Check		OK	03:46
Air Blank		0.000	03:47
Control Test		0.079	03:47
Air Blank		0.000	03:48
Subject Sample	# 1	0.162	03:48
Breath Volume		1.191	03:48
Air Blank		0.000	03:49
Air Blank		0.000	03:50
Subject Sample	# 2	0.161	03:51
Breath Volume		1.390	03:51
Air Blank		0.000	03:51
Control Test		0.080	03:52
Air Blank		0.000	03:52
Diagnostics Check		OK	03:52

Data Download Date/Time: 02/03/2020 16:18

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUM,ZACKARY,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LAFFERTY,N
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/22/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: ROSS

DATE/TIME OF TEST: 02/01/2020 00:46:03

Diagnostics Check		OK	00:39
Air Blank		0.000	00:40
Control Test		0.081	00:40
Air Blank		0.000	00:41
Subject Sample	# 1	0.214	00:41
Breath Volume		2.144	00:41
Air Blank		0.000	00:42
Air Blank		0.000	00:44
Subject Sample	# 2	0.221	00:44
Breath Volume		2.015	00:44
Air Blank		0.000	00:45
Control Test		0.081	00:45
Air Blank		0.000	00:45
Diagnostics Check		OK	00:46

Data Download Date/Time: 02/05/2020 08:57

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCULLOUGH,SEAN,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HINES,PAUL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/22/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:36

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 02/01/2020 05:09:23

Diagnostics Check		OK	04:57	
Air Blank		0.000	04:57	
Control Test		0.081	04:57	
Air Blank		0.000	04:58	
Subject Sample	# 1	0.193	05:01	Volume Not Met
Breath Volume		0.343	05:01	
Air Blank		0.000	05:02	
Air Blank		0.000	05:03	
Subject Sample	# 2	0.180	05:04	
Breath Volume		1.664	05:04	
Air Blank		0.000	05:05	
Air Blank		0.000	05:07	
Subject Sample	# 3	0.193	05:07	
Breath Volume		3.523	05:07	
Air Blank		0.000	05:08	
Control Test		0.081	05:08	
Air Blank		0.000	05:09	
Diagnostics Check		OK	05:09	

Data Download Date/Time: 02/05/2020 08:57

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,LUIS,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HOBART,CLAYTON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/22/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 02/03/2020 22:40:38

Diagnostics Check		OK	22:29	
Air Blank		0.000	22:29	
Control Test		0.081	22:30	
Air Blank		0.000	22:30	
Subject Sample	# 1		22:33	No Sample Provided
Breath Volume		0.000	22:33	
Air Blank		0.000	22:34	
Air Blank		0.000	22:36	
Subject Sample	# 2		22:39	No Sample Provided
Breath Volume		0.000	22:39	
Air Blank		0.000	22:39	
Control Test		0.081	22:40	
Air Blank		0.000	22:40	
Diagnostics Check		OK	22:40	

Data Download Date/Time: 02/05/2020 08:57

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEISKOPF,MICHAEL,D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CLARK,BRENT
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/22/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:52

SUBJECT'S LAST NAME: WILKES

DATE/TIME OF TEST: 02/04/2020 16:18:34

Diagnostics Check		OK	16:12
Air Blank		0.000	16:13
Control Test		0.082	16:13
Air Blank		0.000	16:14
Subject Sample	# 1	0.000	16:14
Breath Volume		1.542	16:14
Air Blank		0.000	16:14
Air Blank		0.000	16:16
Subject Sample	# 2	0.000	16:17
Breath Volume		1.699	16:17
Air Blank		0.000	16:17
Control Test		0.081	16:17
Air Blank		0.000	16:18
Diagnostics Check		OK	16:18

Data Download Date/Time: 02/05/2020 08:57

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINES,PAUL,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LAMOUR,ZACHARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: PEREZ PIOQUINTO

DATE/TIME OF TEST: 02/05/2020 23:53:11

Diagnostics Check		OK	23:45
Air Blank		0.000	23:46
Control Test		0.081	23:46
Air Blank		0.000	23:47
Subject Sample	# 1	0.187	23:48
Breath Volume		2.031	23:48
Air Blank		0.000	23:49
Air Blank		0.000	23:50
Subject Sample	# 2	0.180	23:51
Breath Volume		2.343	23:51
Air Blank		0.000	23:52
Control Test		0.080	23:52
Air Blank		0.000	23:53
Diagnostics Check		OK	23:53

Data Download Date/Time: 03/10/2020 15:22

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LESPERANCE,DANIEL,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HINES,PAUL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: SILVER

DATE/TIME OF TEST: 02/07/2020 01:03:26

Diagnostics Check		OK	00:52	
Air Blank		0.000	00:52	
Control Test		0.080	00:52	
Air Blank		0.000	00:53	
Subject Sample	# 1		00:56	No Sample Provided
Breath Volume		0.000	00:56	
Air Blank		0.000	00:56	
Air Blank		0.000	00:58	
Subject Sample	# 2		01:02	No Sample Provided
Breath Volume		0.000	01:02	
Air Blank		0.000	01:02	
Control Test		0.080	01:02	
Air Blank		0.000	01:03	
Diagnostics Check		OK	01:03	

Data Download Date/Time: 03/10/2020 15:22

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIRANDA,RICHARD,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: NOLL

DATE/TIME OF TEST: 02/07/2020 04:02:53

Diagnostics Check		OK	03:54	
Air Blank		0.000	03:54	
Control Test		0.081	03:55	
Air Blank		0.000	03:55	
Subject Sample	# 1		03:58	No Sample Provided
Breath Volume		0.000	03:58	
Air Blank		0.000	03:59	
Air Blank		0.000	04:01	
Subject Sample	# 2		04:01	Subject Test Refused
Breath Volume		0.000	04:01	
Air Blank		0.000	04:01	
Control Test		0.081	04:02	
Air Blank		0.000	04:02	
Diagnostics Check		OK	04:02	

Data Download Date/Time: 03/10/2020 15:22

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINES,PAUL,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CRUM,ZACKARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 02/08/2020 00:41:10

Diagnostics Check		OK	00:34
Air Blank		0.000	00:34
Control Test		0.081	00:35
Air Blank		0.000	00:35
Subject Sample	# 1	0.156	00:36
Breath Volume		3.570	00:36
Air Blank		0.000	00:37
Air Blank		0.000	00:39
Subject Sample	# 2	0.155	00:39
Breath Volume		3.441	00:39
Air Blank		0.000	00:40
Control Test		0.082	00:40
Air Blank		0.000	00:41
Diagnostics Check		OK	00:41

Data Download Date/Time: 03/10/2020 15:22

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN HOUTEN,JONATHAN,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LESPERANCE,DANIEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 02/08/2020 02:25:23

Diagnostics Check		OK	02:18
Air Blank		0.000	02:18
Control Test		0.082	02:19
Air Blank		0.000	02:19
Subject Sample	# 1	0.115	02:20
Breath Volume		1.890	02:20
Air Blank		0.000	02:20
Air Blank		0.000	02:22
Subject Sample	# 2	0.116	02:23
Breath Volume		1.328	02:23
Air Blank		0.000	02:24
Control Test		0.082	02:24
Air Blank		0.000	02:25
Diagnostics Check		OK	02:25

Data Download Date/Time: 03/10/2020 15:22

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUM,ZACKARY,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CRUM,ZACKARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: MCMAHON

DATE/TIME OF TEST: 02/08/2020 04:07:32

Diagnostics Check		OK	04:04	
Air Blank		0.000	04:05	
Control Test		0.082	04:05	
Air Blank		0.000	04:06	
Subject Sample	# 1		04:06	Subject Test Refused
Breath Volume		0.000	04:06	
Air Blank		0.000	04:06	
Control Test		0.082	04:06	
Air Blank		0.000	04:07	
Diagnostics Check		OK	04:07	

Data Download Date/Time: 03/10/2020 15:22

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCULLOUGH,SEAN,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CRUM,ZACHARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: DUDLEY

DATE/TIME OF TEST: 02/08/2020 19:30:01

Diagnostics Check		OK	19:23
Air Blank		0.000	19:24
Control Test		0.083	19:24
Air Blank		0.000	19:24
Subject Sample	# 1	0.098	19:25
Breath Volume		2.089	19:25
Air Blank		0.000	19:25
Air Blank		0.000	19:27
Subject Sample	# 2	0.098	19:28
Breath Volume		2.261	19:28
Air Blank		0.000	19:29
Control Test		0.082	19:29
Air Blank		0.000	19:29
Diagnostics Check		OK	19:30

Data Download Date/Time: 03/10/2020 15:22

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUM,ZACKARY,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LESPERANCE,D
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:49

SUBJECT'S LAST NAME: WILKINS

DATE/TIME OF TEST: 02/08/2020 21:18:18

Diagnostics Check		OK	21:09	
Air Blank		0.000	21:10	
Control Test		0.082	21:10	
Air Blank		0.000	21:11	
Subject Sample	# 1		21:14	No Sample Provided
Breath Volume		0.000	21:14	
Air Blank		0.000	21:14	
Air Blank		0.000	21:16	
Subject Sample	# 2		21:16	Subject Test Refused
Breath Volume		0.000	21:16	
Air Blank		0.000	21:17	
Control Test		0.082	21:17	
Air Blank		0.000	21:18	
Diagnostics Check		OK	21:18	

Data Download Date/Time: 03/10/2020 15:22

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LESPERANCE,DANIEL,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCCULLOUGH,SEAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: BIERSACK

DATE/TIME OF TEST: 02/09/2020 04:12:14

Diagnostics Check		OK	04:04
Air Blank		0.000	04:04
Control Test		0.082	04:04
Air Blank		0.000	04:05
Subject Sample	# 1	0.156	04:06
Breath Volume		1.750	04:06
Air Blank		0.000	04:06
Air Blank		0.000	04:08
Subject Sample	# 2	0.161	04:10
Breath Volume		1.171	04:10
Air Blank		0.000	04:11
Control Test		0.082	04:11
Air Blank		0.000	04:12
Diagnostics Check		OK	04:12

Data Download Date/Time: 03/10/2020 15:22

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLINE,NICOLE,S
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CARTER,MATTHEW
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: CALLERY

DATE/TIME OF TEST: 02/09/2020 23:17:13

Diagnostics Check		OK	23:11
Air Blank		0.000	23:11
Control Test		0.080	23:11
Air Blank		0.000	23:12
Subject Sample	# 1	0.175	23:12
Breath Volume		3.941	23:12
Air Blank		0.000	23:13
Air Blank		0.000	23:15
Subject Sample	# 2	0.183	23:15
Breath Volume		4.128	23:15
Air Blank		0.000	23:16
Control Test		0.081	23:16
Air Blank		0.000	23:17
Diagnostics Check		OK	23:17

Data Download Date/Time: 03/10/2020 15:31

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOCKERS,NICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCKENZIE,R
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 09:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:37

SUBJECT'S LAST NAME: SIEBENTHAL

DATE/TIME OF TEST: 02/10/2020 11:01:42

Diagnostics Check		OK	10:57	
Air Blank		0.000	10:57	
Control Test		0.081	10:58	
Air Blank		0.000	10:58	
Subject Sample	# 1		11:00	Subject Test Refused
Breath Volume		0.000	11:00	
Air Blank		0.000	11:00	
Control Test		0.082	11:01	
Air Blank		0.000	11:01	
Diagnostics Check		OK	11:01	

Data Download Date/Time: 03/10/2020 15:31

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,MATTHEW,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CARTER,MATTHEW
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 02/12/2020 00:23:48

Diagnostics Check		OK	00:15
Air Blank		0.000	00:15
Control Test		0.082	00:16
Air Blank		0.000	00:16
Subject Sample	# 1	0.021	00:19
Breath Volume		2.761	00:19
Air Blank		0.000	00:19
Air Blank		0.000	00:21
Subject Sample	# 2	0.021	00:22
Breath Volume		1.996	00:22
Air Blank		0.000	00:22
Control Test		0.081	00:23
Air Blank		0.000	00:23
Diagnostics Check		OK	00:23

Data Download Date/Time: 03/10/2020 15:31

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,MATTHEW,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CARTER,MATTHEW
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 02/13/2020 05:16:24

Diagnostics Check		OK	05:10
Air Blank		0.000	05:10
Control Test		0.082	05:10
Air Blank		0.000	05:11
Subject Sample	# 1	0.080	05:11
Breath Volume		1.437	05:11
Air Blank		0.000	05:12
Air Blank		0.000	05:14
Subject Sample	# 2	0.074	05:14
Breath Volume		1.414	05:14
Air Blank		0.000	05:15
Control Test		0.082	05:15
Air Blank		0.000	05:16
Diagnostics Check		OK	05:16

Data Download Date/Time: 03/10/2020 15:31

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LESPERANCE,DANIEL,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CRUM,ZACHARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: KENSLER

DATE/TIME OF TEST: 02/13/2020 20:18:32

Diagnostics Check		OK	20:10
Air Blank		0.000	20:11
Control Test		0.081	20:11
Air Blank		0.000	20:12
Subject Sample	# 1	0.152	20:13
Breath Volume		1.269	20:13
Air Blank		0.000	20:13
Air Blank		0.000	20:15
Subject Sample	# 2	0.162	20:17
Breath Volume		3.140	20:17
Air Blank		0.000	20:17
Control Test		0.081	20:17
Air Blank		0.000	20:18
Diagnostics Check		OK	20:18

Data Download Date/Time: 03/10/2020 15:31

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINES,PAUL,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: GONZALEZ,ALEX
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:44

SUBJECT'S LAST NAME: TIMPSON

DATE/TIME OF TEST: 02/13/2020 22:42:23

Diagnostics Check		OK	22:32
Air Blank		0.000	22:33
Control Test		0.081	22:33
Air Blank		0.000	22:34
Subject Sample	# 1	0.000	22:36
Breath Volume		1.746	22:36
Air Blank		0.000	22:37
Air Blank		0.000	22:39
Subject Sample	# 2	0.000	22:41
Breath Volume		1.156	22:41
Air Blank		0.000	22:41
Control Test		0.081	22:41
Air Blank		0.000	22:42
Diagnostics Check		OK	22:42

Data Download Date/Time: 03/10/2020 15:31

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LESPERANCE,DANIEL,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCCULLOUGH,SEAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: MADSEN

DATE/TIME OF TEST: 02/14/2020 01:48:52

Diagnostics Check		OK	01:39
Air Blank		0.000	01:40
Control Test		0.081	01:40
Air Blank		0.000	01:41
Subject Sample	# 1	0.156	01:43
Breath Volume		1.578	01:43
Air Blank		0.000	01:44
Air Blank		0.000	01:45
Subject Sample	# 2	0.152	01:47
Breath Volume		1.394	01:47
Air Blank		0.000	01:47
Control Test		0.081	01:48
Air Blank		0.000	01:48
Diagnostics Check		OK	01:48

Data Download Date/Time: 03/10/2020 15:32

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCULLOUGH,SEAN,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HINES,PAUL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: MCCLURE

DATE/TIME OF TEST: 02/14/2020 20:27:08

Diagnostics Check		OK	20:20
Air Blank		0.000	20:21
Control Test		0.081	20:21
Air Blank		0.000	20:21
Subject Sample	# 1	0.172	20:22
Breath Volume		2.324	20:22
Air Blank		0.000	20:23
Air Blank		0.000	20:25
Subject Sample	# 2	0.171	20:25
Breath Volume		2.839	20:25
Air Blank		0.000	20:26
Control Test		0.080	20:26
Air Blank		0.000	20:27
Diagnostics Check		OK	20:27

Data Download Date/Time: 03/10/2020 15:32

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LESPERANCE,DANIEL,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CRUM,ZACHARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: SINGLETARY

DATE/TIME OF TEST: 02/14/2020 22:39:41

Diagnostics Check		OK	22:33
Air Blank		0.000	22:34
Control Test		0.081	22:34
Air Blank		0.000	22:35
Subject Sample	# 1	0.000	22:35
Breath Volume		1.171	22:35
Air Blank		0.000	22:36
Air Blank		0.000	22:38
Subject Sample	# 2	0.000	22:38
Breath Volume		1.210	22:38
Air Blank		0.000	22:38
Control Test		0.082	22:39
Air Blank		0.000	22:39
Diagnostics Check		OK	22:39

Data Download Date/Time: 03/10/2020 15:32

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LESPERANCE,DANIEL,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCCULLOUGH,SEAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: MCGRAW

DATE/TIME OF TEST: 02/15/2020 04:32:24

Diagnostics Check		OK	04:21	
Air Blank		0.000	04:21	
Control Test		0.081	04:21	
Air Blank		0.000	04:22	
Subject Sample	# 1		04:25	No Sample Provided
Breath Volume		0.000	04:25	
Air Blank		0.000	04:25	
Air Blank		0.000	04:27	
Subject Sample	# 2		04:31	No Sample Provided
Breath Volume		0.000	04:31	
Air Blank		0.000	04:31	
Control Test		0.081	04:31	
Air Blank		0.000	04:32	
Diagnostics Check		OK	04:32	

Data Download Date/Time: 03/10/2020 15:32

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEWITT,AMY,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HAUSAM,M
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: L521552772511

DATE/TIME OF TEST: 02/16/2020 04:29:50

Diagnostics Check		OK	04:26	
Air Blank		0.000	04:26	
Control Test		0.081	04:27	
Air Blank		0.000	04:27	
Subject Sample	# 1		04:28	Subject Test Refused
Breath Volume		0.000	04:28	
Air Blank		0.000	04:28	
Control Test		0.081	04:29	
Air Blank		0.000	04:29	
Diagnostics Check		OK	04:29	

Data Download Date/Time: 03/10/2020 15:32

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCULLOUGH,SEAN,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HINES,PAUL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: WASHINGTON

DATE/TIME OF TEST: 02/20/2020 22:23:40

Diagnostics Check		OK	22:16
Air Blank		0.000	22:17
Control Test		0.081	22:17
Air Blank		0.000	22:17
Subject Sample	# 1	0.000	22:19
Breath Volume		3.226	22:19
Air Blank		0.000	22:19
Air Blank		0.000	22:21
Subject Sample	# 2	0.000	22:22
Breath Volume		2.531	22:22
Air Blank		0.000	22:22
Control Test		0.081	22:23
Air Blank		0.000	22:23
Diagnostics Check		OK	22:23

Data Download Date/Time: 03/10/2020 15:32

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAMOUR,ZACHARY,T
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCCULLOUGH,SEAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 02/21/2020 00:30:20

Diagnostics Check		OK	00:23
Air Blank		0.000	00:23
Control Test		0.081	00:23
Air Blank		0.000	00:24
Subject Sample	# 1	0.136	00:25
Breath Volume		1.472	00:25
Air Blank		0.000	00:26
Air Blank		0.000	00:28
Subject Sample	# 2	0.133	00:28
Breath Volume		1.308	00:28
Air Blank		0.000	00:29
Control Test		0.079	00:29
Air Blank		0.000	00:30
Diagnostics Check		OK	00:30

Data Download Date/Time: 03/10/2020 15:32

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAMOUR,ZACHARY,T
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HINES,PAUL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: MAXWELL

DATE/TIME OF TEST: 02/21/2020 02:43:18

Diagnostics Check		OK	02:40	
Air Blank		0.000	02:40	
Control Test		0.081	02:41	
Air Blank		0.000	02:41	
Subject Sample	# 1		02:41	Subject Test Refused
Breath Volume		0.000	02:41	
Air Blank		0.000	02:42	
Control Test		0.081	02:42	
Air Blank		0.000	02:43	
Diagnostics Check		OK	02:43	

Data Download Date/Time: 03/10/2020 15:32

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUM,ZACKARY,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: AVEY,J
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: CORDERO

DATE/TIME OF TEST: 02/22/2020 00:05:27

Diagnostics Check		OK	23:56
Air Blank		0.000	23:56
Control Test		0.082	23:57
Air Blank		0.000	23:57
Subject Sample	# 1	0.170	00:00
Breath Volume		1.765	00:00
Air Blank		0.000	00:01
Air Blank		0.000	00:03
Subject Sample	# 2	0.167	00:04
Breath Volume		2.425	00:04
Air Blank		0.000	00:04
Control Test		0.080	00:04
Air Blank		0.000	00:05
Diagnostics Check		OK	00:05

Data Download Date/Time: 03/10/2020 15:32

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCULLOUGH,SEAN,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HINES,PAUL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 02/22/2020 01:43:29

Diagnostics Check		OK	01:32	
Air Blank		0.000	01:32	
Control Test		0.082	01:32	
Air Blank		0.000	01:33	
Subject Sample	# 1		01:36	No Sample Provided
Breath Volume		0.000	01:36	
Air Blank		0.000	01:36	
Air Blank		0.000	01:38	
Subject Sample	# 2		01:42	No Sample Provided
Breath Volume		0.000	01:42	
Air Blank		0.000	01:42	
Control Test		0.082	01:42	
Air Blank		0.000	01:43	
Diagnostics Check		OK	01:43	

Data Download Date/Time: 03/10/2020 15:32

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINES,PAUL,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LAMOUR,ZACHARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: BARGANIER

DATE/TIME OF TEST: 02/22/2020 02:49:23

Diagnostics Check		OK	02:38	
Air Blank		0.000	02:38	
Control Test		0.083	02:38	
Air Blank		0.000	02:39	
Subject Sample	# 1		02:42	No Sample Provided
Breath Volume		0.000	02:42	
Air Blank		0.000	02:42	
Air Blank		0.000	02:44	
Subject Sample	# 2		02:48	No Sample Provided
Breath Volume		0.000	02:48	
Air Blank		0.000	02:48	
Control Test		0.083	02:48	
Air Blank		0.000	02:49	
Diagnostics Check		OK	02:49	

Data Download Date/Time: 03/10/2020 15:32

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,LUIS,A
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BUTTERFIELD,KELLY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: MURRAY

DATE/TIME OF TEST: 02/22/2020 22:47:16

Diagnostics Check		OK	22:35	
Air Blank		0.000	22:36	
Control Test		0.082	22:36	
Air Blank		0.000	22:37	
Subject Sample	# 1		22:40	No Sample Provided
Breath Volume		0.000	22:40	
Air Blank		0.000	22:40	
Air Blank		0.000	22:42	
Subject Sample	# 2		22:45	No Sample Provided
Breath Volume		0.000	22:45	
Air Blank		0.000	22:46	
Control Test		0.082	22:46	
Air Blank		0.000	22:47	
Diagnostics Check		OK	22:47	

Data Download Date/Time: 03/10/2020 15:32

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARRETT,DAVID,S
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CRUM,Z
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: BUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: WOOLEY

DATE/TIME OF TEST: 02/23/2020 04:08:01

Diagnostics Check		OK	04:04	
Air Blank		0.000	04:04	
Control Test		0.082	04:05	
Air Blank		0.000	04:05	
Subject Sample	# 1		04:06	Subject Test Refused
Breath Volume		0.000	04:06	
Air Blank		0.000	04:07	
Control Test		0.081	04:07	
Air Blank		0.000	04:07	
Diagnostics Check		OK	04:08	

Data Download Date/Time: 03/10/2020 15:32

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINES,PAUL,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LAMOUR,ZACHARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: PITTMAN

DATE/TIME OF TEST: 02/27/2020 02:34:41

Diagnostics Check		OK	02:23	
Air Blank		0.000	02:23	
Control Test		0.081	02:24	
Air Blank		0.000	02:24	
Subject Sample	# 1		02:27	No Sample Provided
Breath Volume		0.000	02:27	
Air Blank		0.000	02:28	
Air Blank		0.000	02:30	
Subject Sample	# 2		02:33	No Sample Provided
Breath Volume		0.000	02:33	
Air Blank		0.000	02:33	
Control Test		0.082	02:34	
Air Blank		0.000	02:34	
Diagnostics Check		OK	02:34	

Data Download Date/Time: 03/10/2020 15:32

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUM,ZACKARY,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LAMOUR,Z
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 02/27/2020 23:08:23

Diagnostics Check		OK	23:01
Air Blank		0.000	23:02
Control Test		0.082	23:02
Air Blank		0.000	23:02
Subject Sample	# 1	0.077	23:03
Breath Volume		1.699	23:03
Air Blank		0.000	23:04
Air Blank		0.000	23:06
Subject Sample	# 2	0.079	23:06
Breath Volume		1.425	23:06
Air Blank		0.000	23:07
Control Test		0.082	23:07
Air Blank		0.000	23:08
Diagnostics Check		OK	23:08

Data Download Date/Time: 03/10/2020 15:32

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINES,PAUL,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCCULLOUGH,SEAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: MCCONNELL

DATE/TIME OF TEST: 02/28/2020 00:19:51

Diagnostics Check		OK	00:13
Air Blank		0.000	00:13
Control Test		0.082	00:14
Air Blank		0.000	00:14
Subject Sample	# 1	0.202	00:15
Breath Volume		3.585	00:15
Air Blank		0.000	00:15
Air Blank		0.000	00:17
Subject Sample	# 2	0.202	00:18
Breath Volume		3.324	00:18
Air Blank		0.000	00:18
Control Test		0.081	00:19
Air Blank		0.000	00:19
Diagnostics Check		OK	00:19

Data Download Date/Time: 03/10/2020 15:33

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAMOUR,ZACHARY,T
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HINES,PAUL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: ROMAN-NIEVES

DATE/TIME OF TEST: 02/28/2020 02:04:29

Diagnostics Check		OK	01:54	
Air Blank		0.000	01:54	
Control Test		0.082	01:54	
Air Blank		0.000	01:55	
Subject Sample	# 1		01:58	No Sample Provided
Breath Volume		0.000	01:58	
Air Blank		0.000	01:59	
Air Blank		0.000	02:01	
Subject Sample	# 2		02:03	Subject Test Refused
Breath Volume		0.000	02:03	
Air Blank		0.000	02:03	
Control Test		0.082	02:03	
Air Blank		0.000	02:04	
Diagnostics Check		OK	02:04	

Data Download Date/Time: 03/10/2020 15:33

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARRETT,DAVID,S
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LAFFERTY,NAOMI
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: C150790752250

DATE/TIME OF TEST: 02/29/2020 00:55:00

Diagnostics Check		OK	00:48
Air Blank		0.000	00:48
Control Test		0.082	00:49
Air Blank		0.000	00:49
Subject Sample	# 1	0.240	00:50
Breath Volume		2.687	00:50
Air Blank		0.000	00:51
Air Blank		0.000	00:53
Subject Sample	# 2	0.227	00:53
Breath Volume		2.890	00:53
Air Blank		0.000	00:54
Control Test		0.080	00:54
Air Blank		0.000	00:54
Diagnostics Check		OK	00:55

Data Download Date/Time: 03/10/2020 15:33

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 04:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 02/01/2020 05:11:53

Diagnostics Check		OK	05:05
Air Blank		0.000	05:06
Control Test		0.082	05:06
Air Blank		0.000	05:06
Subject Sample	# 1	0.182	05:07
Breath Volume		4.078	05:07
Air Blank		0.000	05:08
Air Blank		0.000	05:09
Subject Sample	# 2	0.173	05:10
Breath Volume		4.105	05:10
Air Blank		0.000	05:10
Control Test		0.081	05:11
Air Blank		0.000	05:11
Diagnostics Check		OK	05:11

Data Download Date/Time: 02/25/2020 13:01

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 21:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: OCHOA

DATE/TIME OF TEST: 02/01/2020 22:06:56

Diagnostics Check		OK	22:00
Air Blank		0.000	22:01
Control Test		0.081	22:01
Air Blank		0.000	22:01
Subject Sample	# 1	0.115	22:02
Breath Volume		1.714	22:02
Air Blank		0.000	22:03
Air Blank		0.000	22:04
Subject Sample	# 2	0.111	22:05
Breath Volume		1.531	22:05
Air Blank		0.000	22:06
Control Test		0.080	22:06
Air Blank		0.000	22:06
Diagnostics Check		OK	22:06

Data Download Date/Time: 02/25/2020 13:01

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,MASON,E
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: THOMPSON,MASON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: RACZAK

DATE/TIME OF TEST: 02/03/2020 01:18:29

Diagnostics Check		OK	01:12
Air Blank		0.000	01:12
Control Test		0.081	01:12
Air Blank		0.000	01:13
Subject Sample	# 1	0.197	01:13
Breath Volume		2.812	01:13
Air Blank		0.000	01:14
Air Blank		0.000	01:16
Subject Sample	# 2	0.200	01:16
Breath Volume		2.792	01:16
Air Blank		0.000	01:17
Control Test		0.081	01:17
Air Blank		0.000	01:18
Diagnostics Check		OK	01:18

Data Download Date/Time: 02/25/2020 13:01

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,MASON,E
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: THOMPSON,MASON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: YI

DATE/TIME OF TEST: 02/05/2020 03:54:48

Diagnostics Check		OK	03:47
Air Blank		0.000	03:47
Control Test		0.081	03:48
Air Blank		0.000	03:48
Subject Sample	# 1	0.075	03:49
Breath Volume		1.355	03:49
Air Blank		0.000	03:49
Air Blank		0.000	03:51
Subject Sample	# 2	0.086	03:53
Breath Volume		2.285	03:53
Air Blank		0.000	03:53
Control Test		0.080	03:54
Air Blank		0.000	03:54
Diagnostics Check		OK	03:54

Data Download Date/Time: 02/25/2020 13:01

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,JOHNATHAN,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: RUSSELL,DONALD
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 19:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: FLYNN

DATE/TIME OF TEST: 02/11/2020 20:14:26

Diagnostics Check		OK	20:11	
Air Blank		0.000	20:11	
Control Test		0.081	20:11	
Air Blank		0.000	20:12	
Subject Sample	# 1		20:13	Subject Test Refused
Breath Volume		0.000	20:13	
Air Blank		0.000	20:13	
Control Test		0.081	20:13	
Air Blank		0.000	20:14	
Diagnostics Check		OK	20:14	

Data Download Date/Time: 02/25/2020 13:01

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 02/17/2020 02:35:59

Diagnostics Check		OK	02:29
Air Blank		0.000	02:29
Control Test		0.081	02:29
Air Blank		0.000	02:30
Subject Sample	# 1	0.241	02:31
Breath Volume		3.621	02:31
Air Blank		0.000	02:32
Air Blank		0.000	02:33
Subject Sample	# 2	0.227	02:34
Breath Volume		2.425	02:34
Air Blank		0.000	02:35
Control Test		0.081	02:35
Air Blank		0.000	02:35
Diagnostics Check		OK	02:35

Data Download Date/Time: 02/25/2020 13:01

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,MASON,E
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: CAMDEN,DENNIS
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 04:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:52

SUBJECT'S LAST NAME: BERRY

DATE/TIME OF TEST: 02/22/2020 05:16:21

Diagnostics Check		OK	05:13	
Air Blank		0.000	05:13	
Control Test		0.082	05:14	
Air Blank		0.000	05:14	
Subject Sample	# 1		05:14	Subject Test Refused
Breath Volume		0.000	05:14	
Air Blank		0.000	05:15	
Control Test		0.083	05:15	
Air Blank		0.000	05:16	
Diagnostics Check		OK	05:16	

Data Download Date/Time: 02/25/2020 13:01

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001714

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,TERRELL,G
BREATH TEST OPERATOR AGENCY: U of Fl PD
ARRESTING OFFICER: WILLIAMS,TERRELL
ARRESTING OFFICER AGENCY: U of Fl PD
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 09:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:20

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 02/04/2020 10:55:29

Diagnostics Check	OK	10:54	
Air Blank		10:55	RFI Detect
Air Blank	0.000	10:55	

Data Download Date/Time: 03/16/2020 02:05

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001714

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,TERRELL,G
BREATH TEST OPERATOR AGENCY: U of Fl PD
ARRESTING OFFICER: WILLIAMS,TERRELL
ARRESTING OFFICER AGENCY: U of Fl PD
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 09:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:20

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 02/04/2020 11:07:40

Diagnostics Check		OK	10:57	
Air Blank		0.000	10:58	
Control Test		0.077	10:58	
Air Blank		0.000	10:59	
Subject Sample	# 1		11:02	No Sample Provided
Breath Volume		0.000	11:02	
Air Blank		0.000	11:02	
Air Blank		0.000	11:04	
Subject Sample	# 2		11:06	Subject Test Refused
Breath Volume		0.000	11:06	
Air Blank		0.000	11:06	
Control Test		0.078	11:07	
Air Blank		0.000	11:07	
Diagnostics Check		OK	11:07	

Data Download Date/Time: 03/16/2020 02:05

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DONALDSON,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: DONALDSON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: CABALLERO SORIANO

DATE/TIME OF TEST: 02/01/2020 02:43:18

Diagnostics Check		OK	02:31	
Air Blank		0.000	02:32	
Control Test		0.080	02:32	
Air Blank		0.000	02:32	
Subject Sample	# 1	0.088	02:36	Volume Not Met
Breath Volume		0.593	02:36	
Air Blank		0.000	02:36	
Air Blank		0.000	02:38	
Subject Sample	# 2	0.000	02:41	Volume Not Met
Breath Volume		0.109	02:41	
Air Blank		0.000	02:42	
Control Test		0.079	02:42	
Air Blank		0.000	02:43	
Diagnostics Check		OK	02:43	

Data Download Date/Time: 02/07/2020 18:10

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DONALDSON,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: DONALDSON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: CABALLERO SORIANO

DATE/TIME OF TEST: 02/01/2020 02:53:03

Diagnostics Check		OK	02:46
Air Blank		0.000	02:46
Control Test		0.080	02:46
Air Blank		0.000	02:47
Subject Sample	# 1	0.106	02:48
Breath Volume		1.167	02:48
Air Blank		0.000	02:49
Air Blank		0.000	02:50
Subject Sample	# 2	0.107	02:51
Breath Volume		1.929	02:51
Air Blank		0.000	02:52
Control Test		0.079	02:52
Air Blank		0.000	02:52
Diagnostics Check		OK	02:53

Data Download Date/Time: 02/07/2020 18:10

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,DAVID,T
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: MEJIA,DANIEL
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: ROJAS

DATE/TIME OF TEST: 02/02/2020 04:56:01

Diagnostics Check		OK	04:49
Air Blank		0.000	04:49
Control Test		0.080	04:50
Air Blank		0.000	04:50
Subject Sample	# 1	0.080	04:51
Breath Volume		3.027	04:51
Air Blank		0.000	04:52
Air Blank		0.000	04:53
Subject Sample	# 2	0.078	04:54
Breath Volume		3.484	04:54
Air Blank		0.000	04:55
Control Test		0.077	04:55
Air Blank		0.000	04:55
Diagnostics Check		OK	04:56

Data Download Date/Time: 02/07/2020 18:10

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,BRIAN,G
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: GARCIA,CORINA
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: HOIST

DATE/TIME OF TEST: 02/05/2020 05:44:18

Diagnostics Check		OK	05:37
Air Blank		0.000	05:38
Control Test		0.081	05:38
Air Blank		0.000	05:39
Subject Sample	# 1	0.142	05:39
Breath Volume		2.613	05:39
Air Blank		0.000	05:40
Air Blank		0.000	05:42
Subject Sample	# 2	0.138	05:42
Breath Volume		3.082	05:42
Air Blank		0.000	05:43
Control Test		0.081	05:43
Air Blank		0.000	05:44
Diagnostics Check		OK	05:44

Data Download Date/Time: 02/07/2020 18:10

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,DAVID,T
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: TORO,DAVID
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: RATCLIFFE

DATE/TIME OF TEST: 02/07/2020 04:28:44

Diagnostics Check		OK	04:15	
Air Blank		0.000	04:16	
Control Test		0.081	04:16	
Air Blank		0.000	04:17	
Subject Sample	# 1	0.128	04:20	Volume Not Met
Breath Volume		0.203	04:20	
Air Blank		0.000	04:21	
Air Blank		0.000	04:22	
Subject Sample	# 2	0.147	04:23	
Breath Volume		2.214	04:23	
Air Blank		0.000	04:24	
Air Blank		0.000	04:26	
Subject Sample	# 3	0.155	04:27	
Breath Volume		1.824	04:27	
Air Blank		0.000	04:27	
Control Test		0.080	04:28	
Air Blank		0.000	04:28	
Diagnostics Check		OK	04:28	

Data Download Date/Time: 02/07/2020 18:10

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.